



INSTRUMENT PROCESSING SHEET

Agency Hillsborough County Sheriff's Office

S/N 80-007481

Florida Department of
Law Enforcement

Date In 04/25/2025 DI Completion Date 05/20/2025

☐ Ship ☒ P/U ☐ H/D ☐ CMI ☐ EE

Intake	By TDG	Date 05/19/2025	Quality Checks	By TDG	Date 05/19/2025	Flow Calibration	By TDG	Date 05/19/2025																														
☒ Annual ☐ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☐ Power cord ☐ Printer Cable ☒ Static Bag ☐ 12V DC Cable Notes: Dropped off with box.			☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value 107 ☒ Flow Verification (L/s) Flow Column # ATP104 32 mm 0.105* (.139 - .169) 36 mm 0.121* (.156 - .190) 53 mm 0.195* (.228 - .278) 103 mm 0.484 (.447 - .547) ☒ Barometric Pressure Check Gauge ID # 28199 ☒ Stability Checks			Flow Column # ATP106 ☒ 5L/min - 17mm ☒ 15L/min - 53mm ☒ 30L/min - 103mm ☒ R-Value 109 ☒ Post Calibration Verification (L/s) Flow Column # ATP104 32 mm 0.140 (.139 - .169) 36 mm 0.160 (.156 - .190) 53 mm 0.234 (.228 - .278) 103 mm 0.503 (.447 - .547)																																
			<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #/Exp</th></tr></thead><tbody><tr><td>0.050</td><td>MP6286</td><td>202406K 06/19/2026</td></tr><tr><td>0.080</td><td>MP6287</td><td>202406L 06/19/2026</td></tr><tr><td>0.200</td><td>MP6288</td><td>202406N 06/20/2026</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG429602 10/22/2026</td></tr></tbody></table>			Simulator	Serial #	Lot #/Exp	0.050	MP6286	202406K 06/19/2026	0.080	MP6287	202406L 06/19/2026	0.200	MP6288	202406N 06/20/2026	0.080 DGS	N/A	AG429602 10/22/2026	<table border="1"><thead><tr><th colspan="2">Maintenance By</th><th>Date</th></tr></thead><tbody><tr><td>☐ Battery Replacement</td><td></td><td></td></tr><tr><td>☐ Dry Gas Regulator Replacement</td><td></td><td></td></tr><tr><td>☐ Breath Tube Replacement</td><td></td><td></td></tr><tr><td>☐ Other</td><td></td><td></td></tr></tbody></table>			Maintenance By		Date	☐ Battery Replacement			☐ Dry Gas Regulator Replacement			☐ Breath Tube Replacement			☐ Other		
Simulator	Serial #	Lot #/Exp																																				
0.050	MP6286	202406K 06/19/2026																																				
0.080	MP6287	202406L 06/19/2026																																				
0.200	MP6288	202406N 06/20/2026																																				
0.080 DGS	N/A	AG429602 10/22/2026																																				
Maintenance By		Date																																				
☐ Battery Replacement																																						
☐ Dry Gas Regulator Replacement																																						
☐ Breath Tube Replacement																																						
☐ Other																																						
Calibration Adjustment By			Department Inspection By TDG																																			
Barometric Pressure Gauge ID #			Barometric Pressure ID# 28199																																			
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.000</td><td></td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td></td><td></td><td></td></tr><tr><td>0.100</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.300</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>			Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Gauge 1016 Instrument 1012 Mouth Alcohol Solution Lot # 2024-A Acetone Stock Solution Lot # 2023-B							
Simulator	Serial #	Lot #	Expiration																																			
0.000		N/A	N/A																																			
0.040																																						
0.100																																						
0.200																																						
0.300																																						
0.080 DGS	N/A																																					
☐ Post Calibration Adjustment Stability Checks			<table border="1"><thead><tr><th>Simulator</th><th>Serial Number</th></tr></thead><tbody><tr><td>0.000</td><td>MP6284</td></tr><tr><td>Interferent</td><td>MP6285</td></tr><tr><td>0.050</td><td>MP6286</td></tr><tr><td>0.080</td><td>MP6287</td></tr><tr><td>0.200</td><td>MP6288</td></tr></tbody></table>			Simulator	Serial Number	0.000	MP6284	Interferent	MP6285	0.050	MP6286	0.080	MP6287	0.200	MP6288																					
Simulator	Serial Number																																					
0.000	MP6284																																					
Interferent	MP6285																																					
0.050	MP6286																																					
0.080	MP6287																																					
0.200	MP6288																																					
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.050</td><td></td><td></td><td></td></tr><tr><td>0.080</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>			Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			Attachments ☒ Form 41 ☐ Post-Stability Checks ☒ Stability Checks ☒ Flow Calibration ☒ Calibration Certificate ☐ Form 40 ☐ Calibration Adjustment ☐ Other															
Simulator	Serial #	Lot #	Expiration																																			
0.050																																						
0.080																																						
0.200																																						
0.080 DGS	N/A																																					
Notes/Suggested Service: *Outside nominal. (TDG 5/19/25) Discussed low R-value with Agency Inspector Jenny Hammond. Instrument will be sent early next year for Department Inspection to ensure R-value does not fall below 100. (TDG 5/23/25)			☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use																																			
			Shayla Platt Digitally signed by Shayla Platt Date: 2025.05.31 15:39:42 -04'00'																																			
			Phil Nicodemo Digitally signed by Phil Nicodemo Date: 2025.05.29 12:24:01 -04'00'																																			
			Tech Review / Date Admin Review / Date																																			

Flow Calibration Adjustment(s)

Root Cause Analysis Performed Prior? ☒

Performed by TDG

HILLSBOROUGH CSO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-007481
05/19/2025
Software: 8100.27

* Flow Rate Calibration*****

1: Rate (Liters/min) = 5

SQRT(Diff) = 5.742

2: Rate (Liters/min) = 15

SQRT(Diff) = 10.723

3: Rate (Liters/min) = 30

SQRT(Diff) = 20.539

* Dependent Data Scale Factor = 100000 L/min

Independent Data Scale Factor = 256

Rounded Slope = 651

Rounded Intercept = -389523

Correlation = 0.99744

Stability Checks

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L																																																																																																																																																
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083																																																																																																																																																
✓ HILLSBOROUGH CSO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-007481 05/19/2025 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>13:47</td></tr> <tr><td>Control Test</td><td>0.049</td><td>13:48</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:48</td></tr> <tr><td>Control Test</td><td>0.049</td><td>13:49</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:50</td></tr> <tr><td>Control Test</td><td>0.049</td><td>13:50</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:51</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0490</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table> Operator's Signature 	Test	g/210L	Time	Air Blank	0.000	13:47	Control Test	0.049	13:48	Air Blank	0.000	13:48	Control Test	0.049	13:49	Air Blank	0.000	13:50	Control Test	0.049	13:50	Air Blank	0.000	13:51	Control Test Stats			Average	0.0490		Std Dev	0.0000		Rel Std Dev(%)	0.0000		✓ HILLSBOROUGH CSO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-007481 05/19/2025 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>13:59</td></tr> <tr><td>Control Test</td><td>0.080</td><td>13:59</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:00</td></tr> <tr><td>Control Test</td><td>0.080</td><td>14:01</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:01</td></tr> <tr><td>Control Test</td><td>0.080</td><td>14:02</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:02</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0800</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table> Operator's Signature 	Test	g/210L	Time	Air Blank	0.000	13:59	Control Test	0.080	13:59	Air Blank	0.000	14:00	Control Test	0.080	14:01	Air Blank	0.000	14:01	Control Test	0.080	14:02	Air Blank	0.000	14:02	Control Test Stats			Average	0.0800		Std Dev	0.0000		Rel Std Dev(%)	0.0000		✓ HILLSBOROUGH CSO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-007481 05/19/2025 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>13:38</td></tr> <tr><td>Control Test</td><td>0.198</td><td>13:39</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:39</td></tr> <tr><td>Control Test</td><td>0.198</td><td>13:40</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:41</td></tr> <tr><td>Control Test</td><td>0.198</td><td>13:41</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:42</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.1980</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table> Operator's Signature 	Test	g/210L	Time	Air Blank	0.000	13:38	Control Test	0.198	13:39	Air Blank	0.000	13:39	Control Test	0.198	13:40	Air Blank	0.000	13:41	Control Test	0.198	13:41	Air Blank	0.000	13:42	Control Test Stats			Average	0.1980		Std Dev	0.0000		Rel Std Dev(%)	0.0000		✓ HILLSBOROUGH CSO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-007481 05/19/2025 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>14:05</td></tr> <tr><td>Control Test</td><td>0.081</td><td>14:05</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:05</td></tr> <tr><td>Control Test</td><td>0.080</td><td>14:06</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:06</td></tr> <tr><td>Control Test</td><td>0.081</td><td>14:06</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:07</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0807</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7157</td><td></td></tr> </tbody> </table> Operator's Signature _____	Test	g/210L	Time	Air Blank	0.000	14:05	Control Test	0.081	14:05	Air Blank	0.000	14:05	Control Test	0.080	14:06	Air Blank	0.000	14:06	Control Test	0.081	14:06	Air Blank	0.000	14:07	Control Test Stats			Average	0.0807		Std Dev	0.0006		Rel Std Dev(%)	0.7157	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	13:47																																																																																																																																																	
Control Test	0.049	13:48																																																																																																																																																	
Air Blank	0.000	13:48																																																																																																																																																	
Control Test	0.049	13:49																																																																																																																																																	
Air Blank	0.000	13:50																																																																																																																																																	
Control Test	0.049	13:50																																																																																																																																																	
Air Blank	0.000	13:51																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0490																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	13:59																																																																																																																																																	
Control Test	0.080	13:59																																																																																																																																																	
Air Blank	0.000	14:00																																																																																																																																																	
Control Test	0.080	14:01																																																																																																																																																	
Air Blank	0.000	14:01																																																																																																																																																	
Control Test	0.080	14:02																																																																																																																																																	
Air Blank	0.000	14:02																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0800																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	13:38																																																																																																																																																	
Control Test	0.198	13:39																																																																																																																																																	
Air Blank	0.000	13:39																																																																																																																																																	
Control Test	0.198	13:40																																																																																																																																																	
Air Blank	0.000	13:41																																																																																																																																																	
Control Test	0.198	13:41																																																																																																																																																	
Air Blank	0.000	13:42																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.1980																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	14:05																																																																																																																																																	
Control Test	0.081	14:05																																																																																																																																																	
Air Blank	0.000	14:05																																																																																																																																																	
Control Test	0.080	14:06																																																																																																																																																	
Air Blank	0.000	14:06																																																																																																																																																	
Control Test	0.081	14:06																																																																																																																																																	
Air Blank	0.000	14:07																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0807																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7157																																																																																																																																																		

Florida Department of Law Enforcement

Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: HILLSBOROUGH CSO
Time of Inspection: 13:07

Date of Inspection: 05/20/2025

Serial Number: 80-007481
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202406K Exp: 06/19/2026	0.08g/210L Test (g/210L) Lot#:202406L Exp: 06/19/2026	0.20g/210L Test (g/210L) Lot#:202406N Exp: 06/20/2026	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG429602 Exp: 10/22/2026
0.000	0.048	0.080	0.198	0.082
0.000	0.048	0.079	0.199	0.081
0.000	0.048	0.078	0.199	0.082
0.000	0.049	0.078	0.199	0.081
0.000	0.049	0.079	0.198	0.081
0.000	0.049	0.079	0.198	0.081
0.000	0.049	0.079	0.198	0.081
0.000	0.049	0.078	0.198	0.081
0.000	0.049	0.078	0.198	0.081
0.000	0.049	0.079	0.198	0.081


Standard Deviations	0.0004	0.0006	0.0004	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



TAYLOR D GUTSCHOW

Signature and Printed Name

05/20/2025
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-007481, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-007481</u>	UNCERTAINTY* $\pm$
Owning Agency:	<u>HILLSBOROUGH CSO</u>	0.050 g/ 210 L 0.004
Calibration Date:	<u>05/20/2025</u>	0.080 g/ 210 L 0.004
Calibration Time:	<u>13:07</u>	0.200 g/ 210 L 0.007
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.
This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

Taylor
Gutschow
Digitally signed by Taylor
Gutschow
Date: 2025.05.23 11:13:24
-04'00'

05/20/2025
Date
TAYLOR D GUTSCHOW,
Department Inspector

FDLE/ATP Form 69 December 2021
Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality