



INSTRUMENT PROCESSING SHEET

Agency Highlands County Sheriff's OfficeS/N 80-007507Florida Department of
Law EnforcementDate In 06/25/2025 DI Completion Date 07/17/2025☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake	By	TDG	Date	07/14/2025	Quality Checks	By	TDG	Date	07/15/2025	Flow Calibration	By	Date																																							
☒ Annual ☐ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☒ Power cord ☐ Printer Cable ☒ Static Bag ☐ 12V DC Cable Notes: <u>Agency Inspector Abeln reports instrument failed its last inspection at 0.080 ARS Test.</u>					☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>222</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP104</u> 32 mm <u>0.152</u> (.139 - .169) 36 mm <u>0.167</u> (.156 - .190) 53 mm <u>0.234</u> (.228 - .278) 103 mm <u>0.492</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>28199</u> ☒ Stability Checks					Flow Column # _____ ☐ 5L/min – 17mm ☐ 15L/min – 53mm ☐ 30L/min – 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																																									
					<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #/Exp</th></tr></thead><tbody><tr><td>0.050</td><td>MP6286</td><td>202406K 06/19/2026</td></tr><tr><td>0.080</td><td>MP6287</td><td>202406L 06/19/2026</td></tr><tr><td>0.200</td><td>MP6288</td><td>202406N 06/20/2026</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG429602 10/22/2026</td></tr></tbody></table>			Simulator	Serial #	Lot #/Exp	0.050	MP6286	202406K 06/19/2026	0.080	MP6287	202406L 06/19/2026	0.200	MP6288	202406N 06/20/2026	0.080 DGS	N/A	AG429602 10/22/2026	<table border="1"><thead><tr><th colspan="2">Maintenance By _____ Date _____</th></tr><tr><td>☐ Battery Replacement</td><td></td></tr><tr><td>☐ Dry Gas Regulator Replacement</td><td></td></tr><tr><td>☐ Breath Tube Replacement</td><td></td></tr><tr><td>☐ Other _____</td><td></td></tr></thead></table>			Maintenance By _____ Date _____		☐ Battery Replacement		☐ Dry Gas Regulator Replacement		☐ Breath Tube Replacement		☐ Other _____																	
Simulator	Serial #	Lot #/Exp																																																	
0.050	MP6286	202406K 06/19/2026																																																	
0.080	MP6287	202406L 06/19/2026																																																	
0.200	MP6288	202406N 06/20/2026																																																	
0.080 DGS	N/A	AG429602 10/22/2026																																																	
Maintenance By _____ Date _____																																																			
☐ Battery Replacement																																																			
☐ Dry Gas Regulator Replacement																																																			
☐ Breath Tube Replacement																																																			
☐ Other _____																																																			
Calibration Adjustment By _____					Department Inspection By <u>TDG</u>																																														
Barometric Pressure Gauge _____ ID # _____					Barometric Pressure ID# <u>28199</u> Gauge <u>1019</u> Instrument <u>1016</u> Mouth Alcohol Solution Lot # <u>2024-A</u> Acetone Stock Solution Lot # <u>2023-B</u>																																														
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.000</td><td></td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td></td><td></td><td></td></tr><tr><td>0.100</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.300</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>					Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			<table border="1"><thead><tr><th>Simulator</th><th>Serial Number</th></tr></thead><tbody><tr><td>0.000</td><td>MP6284</td></tr><tr><td>Interferent</td><td>MP6285</td></tr><tr><td>0.050</td><td>MP6286</td></tr><tr><td>0.080</td><td>MP6287</td></tr><tr><td>0.200</td><td>MP6288</td></tr></tbody></table>							Simulator	Serial Number	0.000	MP6284	Interferent	MP6285	0.050	MP6286	0.080	MP6287	0.200	MP6288
Simulator	Serial #	Lot #	Expiration																																																
0.000		N/A	N/A																																																
0.040																																																			
0.100																																																			
0.200																																																			
0.300																																																			
0.080 DGS	N/A																																																		
Simulator	Serial Number																																																		
0.000	MP6284																																																		
Interferent	MP6285																																																		
0.050	MP6286																																																		
0.080	MP6287																																																		
0.200	MP6288																																																		
☐ Post Calibration Adjustment Stability Checks					<table border="1"><thead><tr><th colspan="2">Attachments</th></tr></thead><tbody><tr><td>☒ Form 41</td><td>☐ Post-Stability Checks</td></tr><tr><td>☒ Stability Checks</td><td>☐ Flow Calibration</td></tr><tr><td>☒ Calibration Certificate</td><td>☐ Form 40</td></tr><tr><td>☐ Calibration Adjustment</td><td>☐ Other _____</td></tr></tbody></table>							Attachments		☒ Form 41	☐ Post-Stability Checks	☒ Stability Checks	☐ Flow Calibration	☒ Calibration Certificate	☐ Form 40	☐ Calibration Adjustment	☐ Other _____																														
Attachments																																																			
☒ Form 41	☐ Post-Stability Checks																																																		
☒ Stability Checks	☐ Flow Calibration																																																		
☒ Calibration Certificate	☐ Form 40																																																		
☐ Calibration Adjustment	☐ Other _____																																																		
Notes/Suggested Service: _____					☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use																																														
					<table border="1"><tr><td>Phil Nicodemus Digitally signed by Phil Nicodemus Date: 2025.07.18 14:37:50 -0400</td><td>Shayla Platt Digitally signed by Shayla Platt Date: 2025.07.21 13:50:11 -0400</td></tr></table>							Phil Nicodemus Digitally signed by Phil Nicodemus Date: 2025.07.18 14:37:50 -0400	Shayla Platt Digitally signed by Shayla Platt Date: 2025.07.21 13:50:11 -0400																																						
Phil Nicodemus Digitally signed by Phil Nicodemus Date: 2025.07.18 14:37:50 -0400	Shayla Platt Digitally signed by Shayla Platt Date: 2025.07.21 13:50:11 -0400																																																		
					Tech Review / Date _____ Admin Review / Date _____																																														

Stability Checks

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L																																																																																																																																				
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083																																																																																																																																				
✓ HIGHLANDS COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 07/15/2025 Software: 8100.27 Test g/210L Time <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>14:37</td></tr> <tr><td>Control Test</td><td>0.048</td><td>14:38</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:38</td></tr> <tr><td>Control Test</td><td>0.049</td><td>14:39</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:39</td></tr> <tr><td>Control Test</td><td>0.048</td><td>14:40</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:41</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.1483</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel. Std Dev(%)</td><td>1.1945</td><td></td></tr> </table> Operator's Signature 	Air Blank	0.000	14:37	Control Test	0.048	14:38	Air Blank	0.000	14:38	Control Test	0.049	14:39	Air Blank	0.000	14:39	Control Test	0.048	14:40	Air Blank	0.000	14:41	Control Test Stats			Average	0.1483		Std Dev	0.0006		Rel. Std Dev(%)	1.1945		✓ HIGHLANDS COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 07/15/2025 Software: 8100.27 Test g/210L Time <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>14:48</td></tr> <tr><td>Control Test</td><td>0.079</td><td>14:49</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:49</td></tr> <tr><td>Control Test</td><td>0.079</td><td>14:50</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:51</td></tr> <tr><td>Control Test</td><td>0.079</td><td>14:51</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:52</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0790</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel. Std Dev(%)</td><td>0.0000</td><td></td></tr> </table> Operator's Signature 	Air Blank	0.000	14:48	Control Test	0.079	14:49	Air Blank	0.000	14:49	Control Test	0.079	14:50	Air Blank	0.000	14:51	Control Test	0.079	14:51	Air Blank	0.000	14:52	Control Test Stats			Average	0.0790		Std Dev	0.0000		Rel. Std Dev(%)	0.0000		✓ HIGHLANDS COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 07/15/2025 Software: 8100.27 Test g/210L Time <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>14:31</td></tr> <tr><td>Control Test</td><td>0.196</td><td>14:31</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:32</td></tr> <tr><td>Control Test</td><td>0.197</td><td>14:33</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:33</td></tr> <tr><td>Control Test</td><td>0.198</td><td>14:34</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:34</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.1570</td><td></td></tr> <tr><td>Std Dev</td><td>0.0010</td><td></td></tr> <tr><td>Rel. Std Dev(%)</td><td>0.5076</td><td></td></tr> </table> Operator's Signature 	Air Blank	0.000	14:31	Control Test	0.196	14:31	Air Blank	0.000	14:32	Control Test	0.197	14:33	Air Blank	0.000	14:33	Control Test	0.198	14:34	Air Blank	0.000	14:34	Control Test Stats			Average	0.1570		Std Dev	0.0010		Rel. Std Dev(%)	0.5076		✓ HIGHLANDS COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 07/15/2025 Software: 8100.27 Test g/210L Time <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>14:54</td></tr> <tr><td>Control Test</td><td>0.079</td><td>14:55</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:55</td></tr> <tr><td>Control Test</td><td>0.079</td><td>14:55</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:56</td></tr> <tr><td>Control Test</td><td>0.079</td><td>14:56</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:57</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0790</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel. Std Dev(%)</td><td>0.0000</td><td></td></tr> </table> Operator's Signature 	Air Blank	0.000	14:54	Control Test	0.079	14:55	Air Blank	0.000	14:55	Control Test	0.079	14:55	Air Blank	0.000	14:56	Control Test	0.079	14:56	Air Blank	0.000	14:57	Control Test Stats			Average	0.0790		Std Dev	0.0000		Rel. Std Dev(%)	0.0000	
Air Blank	0.000	14:37																																																																																																																																					
Control Test	0.048	14:38																																																																																																																																					
Air Blank	0.000	14:38																																																																																																																																					
Control Test	0.049	14:39																																																																																																																																					
Air Blank	0.000	14:39																																																																																																																																					
Control Test	0.048	14:40																																																																																																																																					
Air Blank	0.000	14:41																																																																																																																																					
Control Test Stats																																																																																																																																							
Average	0.1483																																																																																																																																						
Std Dev	0.0006																																																																																																																																						
Rel. Std Dev(%)	1.1945																																																																																																																																						
Air Blank	0.000	14:48																																																																																																																																					
Control Test	0.079	14:49																																																																																																																																					
Air Blank	0.000	14:49																																																																																																																																					
Control Test	0.079	14:50																																																																																																																																					
Air Blank	0.000	14:51																																																																																																																																					
Control Test	0.079	14:51																																																																																																																																					
Air Blank	0.000	14:52																																																																																																																																					
Control Test Stats																																																																																																																																							
Average	0.0790																																																																																																																																						
Std Dev	0.0000																																																																																																																																						
Rel. Std Dev(%)	0.0000																																																																																																																																						
Air Blank	0.000	14:31																																																																																																																																					
Control Test	0.196	14:31																																																																																																																																					
Air Blank	0.000	14:32																																																																																																																																					
Control Test	0.197	14:33																																																																																																																																					
Air Blank	0.000	14:33																																																																																																																																					
Control Test	0.198	14:34																																																																																																																																					
Air Blank	0.000	14:34																																																																																																																																					
Control Test Stats																																																																																																																																							
Average	0.1570																																																																																																																																						
Std Dev	0.0010																																																																																																																																						
Rel. Std Dev(%)	0.5076																																																																																																																																						
Air Blank	0.000	14:54																																																																																																																																					
Control Test	0.079	14:55																																																																																																																																					
Air Blank	0.000	14:55																																																																																																																																					
Control Test	0.079	14:55																																																																																																																																					
Air Blank	0.000	14:56																																																																																																																																					
Control Test	0.079	14:56																																																																																																																																					
Air Blank	0.000	14:57																																																																																																																																					
Control Test Stats																																																																																																																																							
Average	0.0790																																																																																																																																						
Std Dev	0.0000																																																																																																																																						
Rel. Std Dev(%)	0.0000																																																																																																																																						

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: HIGHLANDS COUNTY SO
Time of Inspection: 12:51

Date of Inspection: 07/17/2025

Serial Number: 80-007507
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202406K Exp: 06/19/2026	0.08g/210L Test (g/210L) Lot#:202406L Exp: 06/19/2026	0.20g/210L Test (g/210L) Lot#:202406N Exp: 06/20/2026	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG429602 Exp: 10/22/2026
0.000	0.049	0.080	0.200	0.079
0.000	0.049	0.080	0.201	0.079
0.000	0.049	0.080	0.200	0.079
0.000	0.049	0.080	0.200	0.079
0.000	0.049	0.080	0.201	0.078
0.000	0.049	0.080	0.201	0.079
0.000	0.049	0.081	0.201	0.079
0.000	0.049	0.080	0.201	0.079
0.000	0.049	0.080	0.200	0.078
0.000	0.049	0.080	0.200	0.079


Standard Deviations	0.0000	0.0003	0.0005	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



TAYLOR D GUTSCHOW

Signature and Printed Name

07/17/2025
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-007507, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-007507</u>	UNCERTAINTY* $\pm$
Owning Agency:	<u>HIGHLANDS COUNTY SO</u>	0.050 g/ 210 L 0.004
Calibration Date:	<u>07/17/2025</u>	0.080 g/ 210 L 0.004
Calibration Time:	<u>12:51</u>	0.200 g/ 210 L 0.007
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

Taylor
Gutschow
Digitally signed by Taylor
Gutschow
Date: 2025.07.17 15:39:59
-04'00'

07/17/2025

Date

TAYLOR D GUTSCHOW,

Department Inspector

FDLE/ATP Form 69 December 2021

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality