

Florida Department of Law Enforcement Alcohol Testing Program

REGISTRATION OF EVIDENTIARY BREATH TEST INSTRUMENT

MANUFACTURER: CMI, Inc.
MODEL: Intoxilyzer 8000
SERIAL NUMBER: 80-006924
OWNER: Broward County Sheriff's Office
DATE OF REGISTRATION: December 12th, 2017

The above instrument is hereby approved for evidentiary breath alcohol testing in the State of Florida pursuant to Chapter 11D-8, Florida Administrative Code. This instrument and related records are subject to inspection at any time by the Florida Department of Law Enforcement.



Authorized Representative
Alcohol Testing Program
Florida Department of Law Enforcement



INSTRUMENT PROCESSING SHEET

Florida Department of
Law Enforcement

Agency Broward County Sheriff's Office

S/N 80-006924

Date In 11/27/2017 DI Completion Date 11/28/2017

☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake Performed By <u>DEER</u> ☒ Annual ☒ Registration ☒ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☐ Power cord ☐ Printer Cable ☒ Static Bag ☐ 12V DC Cable Notes: _____ 	Quality Checks Performed By <u>DEER</u> ☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>164</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP101</u> 32 mm <u>.164</u> (.139 - .169) 36 mm <u>.179</u> (.156 - .190) 53 mm <u>.253</u> (.228 - .278) 103 mm <u>.519</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>28663</u> ☒ Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> <tr> <td>0.050</td> <td>SD3967</td> <td>201707D 07/25/2019</td> </tr> <tr> <td>0.080</td> <td>SD3968</td> <td>201707E 07/25/2019</td> </tr> <tr> <td>0.200</td> <td>SD3969</td> <td>201707C 07/24/2019</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG626604 09/22/2018</td> </tr> </table>	Simulator	Serial #	Lot #/Exp	0.050	SD3967	201707D 07/25/2019	0.080	SD3968	201707E 07/25/2019	0.200	SD3969	201707C 07/24/2019	0.080 DGS	N/A	AG626604 09/22/2018	Flow Calibration Performed By _____ Flow Column # _____ ☐ 5L/min – 17mm ☐ 15L/min – 53mm ☐ 30L/min – 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547) Maintenance Performed By _____ ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☐ Other _____ Temperature Checks Performed By <u>DEER</u> ☒ Lab Temp °C <u>22.01c</u> External Digital Therm. ID#: <u>300918</u> ☒ 34°C +/-2 Serial #: <u>SD3967</u> ☒ 34°C +/-2 Serial #: <u>SD3968</u> ☒ 34°C +/-2 Serial #: <u>SD3969</u>																																	
Simulator	Serial #	Lot #/Exp																																																
0.050	SD3967	201707D 07/25/2019																																																
0.080	SD3968	201707E 07/25/2019																																																
0.200	SD3969	201707C 07/24/2019																																																
0.080 DGS	N/A	AG626604 09/22/2018																																																
Final Release Date FDLE DEC 08 2017 Alcohol Testing Program	Calibration Adjustment Performed By _____ Barometric Pressure Gauge ID # _____ <table border="1" style="width:100%; border-collapse: collapse;"> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.300</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </table> ☐ Post Calibration Adjustment Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> <tr> <td>0.050</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </table>		Simulator	Serial Number	Lot Number	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Simulator	Serial Number	Lot Number	Expiration	0.050				0.080				0.200				0.080 DGS	N/A		
Simulator	Serial Number	Lot Number	Expiration																																															
0.000		N/A	N/A																																															
0.040																																																		
0.100																																																		
0.200																																																		
0.300																																																		
0.080 DGS	N/A																																																	
Simulator	Serial Number	Lot Number	Expiration																																															
0.050																																																		
0.080																																																		
0.200																																																		
0.080 DGS	N/A																																																	
Department Inspection Performed By <u>DEER</u> Barometric Pressure ID# <u>68639</u> Gauge <u>1017</u> Instrument <u>1018</u> Mouth Alcohol Solution Lot # <u>2016-C</u> Acetone Stock Solution Lot # <u>2016-B</u> <table border="1" style="width:100%; border-collapse: collapse;"> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> <tr> <td>0.000</td> <td>SD3965</td> </tr> <tr> <td>Interferent</td> <td>SD3966</td> </tr> <tr> <td>0.050</td> <td>SD3967</td> </tr> <tr> <td>0.080</td> <td>SD3968</td> </tr> <tr> <td>0.200</td> <td>SD3969</td> </tr> </table> Attachments ☒ Form 41 ☐ Flow Calibration ☒ Stability Checks ☐ Form 40 ☐ Calibration Adjustment ☒ Other <u>Form 47</u> ☐ Post-Stability Checks		Simulator	Serial Number	0.000	SD3965	Interferent	SD3966	0.050	SD3967	0.080	SD3968	0.200	SD3969	Notes/Suggested Service: <u>Emailed</u> ☒ APPROVED 																																				
Simulator	Serial Number																																																	
0.000	SD3965																																																	
Interferent	SD3966																																																	
0.050	SD3967																																																	
0.080	SD3968																																																	
0.200	SD3969																																																	
☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use <u>Q99m</u> 12/6/17 Tech Review / Date <u>Jf Deaton</u> 12/8/17 Admin Review / Date																																																		

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: BROWARD COUNTY SO
Time of Inspection: 07:46

Date of Inspection: 11/28/2017

Serial Number: 80-006924
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201707D Exp: 07/25/2019	0.08g/210L Test (g/210L) Lot#:201707E Exp: 07/25/2019	0.20g/210L Test (g/210L) Lot#:201707C Exp: 07/24/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG626604 Exp: 09/22/2018
0.000	0.048	0.080	0.197	0.079
0.000	0.048	0.080	0.199	0.079
0.000	0.048	0.080	0.199	0.079
0.000	0.049	0.080	0.199	0.078
0.000	0.048	0.080	0.199	0.079
0.000	0.050	0.081	0.199	0.078
0.000	0.049	0.081	0.200	0.078
0.000	0.049	0.080	0.199	0.079
0.000	0.049	0.080	0.200	0.079
0.000	0.049	0.080	0.199	0.078

Standard Deviations	0.0006	0.0004	0.0008	0.0005
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0005 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

David E Reyes-Rivera DAVID E REYES-RIVERA
Signature and Printed Name

11/28/2017
Date

TYPE OF TEST	SERIAL NUMBER	AGENCY	DATE	PERFORMED BY
Stabilities	80-006924	Broward County Sheriff's Office	11/28/2017	<i>Jul</i>

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L
SN: SD3967 Temp: 34.07c	SN: SD3968 Temp: 34.01c	SN: SD3969 Temp: 34.06c	Lot AG626604
0.047 to 0.053 ☒	0.077 to 0.083 ☒	0.194 to 0.206 ☒	0.077 to 0.083 ☒
BROWARD COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 11/28/2017 Software: 8100.27 Test g/210L Time Air Blank 0.000 05:53 Control Test 0.047 05:53 Air Blank 0.000 05:54 Control Test 0.047 05:54 Air Blank 0.000 05:55 Control Test 0.046 05:56 Air Blank 0.000 05:56 Control Test Stats Average 0.0473 Std Dev 0.0006 Rel Std Dev(%) 1.2198	BROWARD COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 11/28/2017 Software: 8100.27 Test g/210L Time Air Blank 0.000 05:57 Control Test 0.078 05:58 Air Blank 0.000 05:58 Control Test 0.078 05:59 Air Blank 0.000 06:00 Control Test 0.079 06:00 Air Blank 0.000 06:01 Control Test Stats Average 0.0783 Std Dev 0.0006 Rel Std Dev(%) 0.7370	BROWARD COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 11/28/2017 Software: 8100.27 Test g/210L Time Air Blank 0.000 06:02 Control Test 0.196 06:03 Air Blank 0.000 06:03 Control Test 0.196 06:04 Air Blank 0.000 06:05 Control Test 0.196 06:05 Air Blank 0.000 06:06 Control Test Stats Average 0.1960 Std Dev 0.0000 Rel Std Dev(%) 0.0000	BROWARD COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 11/28/2017 Software: 8100.27 Test g/210L Time Air Blank 0.000 06:07 Control Test 0.079 06:07 Air Blank 0.000 06:08 Control Test 0.079 06:08 Air Blank 0.000 06:09 Control Test 0.079 06:09 Air Blank 0.000 06:10 Control Test Stats Average 0.0790 Std Dev 0.0000 Rel Std Dev(%) 0.0000
<i>Jul</i> Operator's Signature	<i>Jul</i> Operator's Signature	<i>Jul</i> Operator's Signature	<i>Jul</i> Operator's Signature

11/28/17

Return Material Authorization Form

Ship to:

☒ CMI, Inc.

☐ Enforcement Electronics

Shipment to repair facility authorized by: Brett Kirkland on 8/31/2017

Items Returned: Instrument ☒ Supplies ☐ Other ☐ Describe: _____

Instrument Model: Intoxilyzer 8000 Serial Number: 80-006924

Bill To Address:

Ship to Address:

FDLE

ATP Attn: David Reyes-Rivera

4700 Terminal Drive Suite 1

Fort Myers, FL 33907

Reason for Return:

New instrument did not pass stability tests, tried to do an optical bench calibration and
the 0.040 solution failed, I had to redo that solution. After the OBC the instrument again
failed the post calibration tests. Instrument is not registered and has not been sent to the field
This is a warranty repair.

Please choose one of the following options:

- ☐ 1. I _____, authorize all repairs.
- ☐ 2. I _____, authorize repairs up to \$_____.
- ☐ 3. I require an estimate **BEFORE** any repairs will be authorized and/ or conducted.

Please contact: Name: _____

Phone #: _____ Email: _____

ATP Contact Name: _____ ATP Email: _____



Alcohol Testing Program

INSTRUMENT PROCESSING SHEET

Agency Broward County Sheriff's OfficeS/N 80-006924Date In 8/30/2017Date Out 8/31/2017☒ Ship ☐ P/U ☐ H/D ☒ CMI ☐ EE

Intake Performed By <u>hell</u> ☒ Registration ☒ Annual ☐ Return from CMI ☐ Return from Enforcement Electronics ☐ Other _____ Visual Inspection: ☒ Case ☒ Handle ☒ Dry Gas Holder ☒ Feet ☒ Keyboard/Plug ☒ Back/Plugs ☒ Screws tight ☒ Breath Hose Other Equipment: ☒ Power cord ☐ Printer Cable ☒ Other <u>Anti Static Bag</u> Notes: <u>New Instrument</u> 	Quality Checks Performed By <u>hell</u> ☒ Lab Temp °C <u>21.36</u> ☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>159</u> ☒ Flow Verification (L/s) Flow Column # <u>28663</u> 32mm <u>.164</u> (.139 - .169) 36mm <u>.179</u> (.156 - .190) 53mm <u>.250</u> (.228 - .278) 103mm <u>.515</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>28663</u> ☒ Stability Checks <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.05</td> <td>SD3967</td> <td>201603D 03/08/2018</td> </tr> <tr> <td>0.08</td> <td>SD3963</td> <td>201611B 11/15/2018</td> </tr> <tr> <td>0.20</td> <td>SD3969</td> <td>201604C 04/05/2018</td> </tr> <tr> <td>0.08 DGS</td> <td>N/A</td> <td>AG626604 09/22/2018</td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.05	SD3967	201603D 03/08/2018	0.08	SD3963	201611B 11/15/2018	0.20	SD3969	201604C 04/05/2018	0.08 DGS	N/A	AG626604 09/22/2018	Flow Calibration Performed By <u>hell</u> ☒ Flow Calibration N/A ☐ Flow Calibration Complete Flow Column # _____ ☐ 5L/min - <u>37mm</u> ☐ 15L/min - <u>59mm</u> ☐ 30L/min - <u>103mm</u> ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32mm _____ (.139 - .169) 36mm _____ (.156 - .190) 53mm _____ (.228 - .278) 103mm _____ (.447 - .547) Maintenance Performed By _____ ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☐ Other _____ Quality Checks Cont. Performed By <u>hell</u> Simulator Temperatures °C External Digital Therm. ID#: <u>300918</u> ☒ 34°C +/- .2 Serial #: <u>SD3967</u> ☒ 34°C +/- .2 Serial #: <u>SD3963</u> ☒ 34°C +/- .2 Serial #: <u>SD3969</u>
Simulator	Serial #	Lot #/Exp															
0.05	SD3967	201603D 03/08/2018															
0.08	SD3963	201611B 11/15/2018															
0.20	SD3969	201604C 04/05/2018															
0.08 DGS	N/A	AG626604 09/22/2018															

Calibration Adjustment Performed By <u>hell</u> ☐ Calibration Adjustment N/A ☒ Calibration Adjustment Complete Barometric Pressure Gauge <u>1015</u> ID # <u>68639</u> <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>2235</td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td>2236</td> <td>16320</td> <td>10/21/2018</td> </tr> <tr> <td>0.100</td> <td>2237</td> <td>17060</td> <td>02/14/2019</td> </tr> <tr> <td>0.200</td> <td>2238</td> <td>17090</td> <td>02/24/2019</td> </tr> <tr> <td>0.300</td> <td>2239</td> <td>16410</td> <td>12/19/2018</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>34416080A2</td> <td>02/05/2019</td> </tr> </tbody> </table> ☒ Post Calibration Adjustment Stability Checks <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.05</td> <td>SD3967</td> <td>201603D</td> <td>03/08/2018</td> </tr> <tr> <td>0.08</td> <td>SD3963</td> <td>201611B</td> <td>11/15/2018</td> </tr> <tr> <td>0.20</td> <td>SD3969</td> <td>201604C</td> <td>04/05/2018</td> </tr> <tr> <td>0.08 DGS</td> <td>N/A</td> <td>AG626604</td> <td>09/22/2018</td> </tr> </tbody> </table>	Simulator	Serial Number	Lot Number	Expiration	0.000	2235	N/A	N/A	0.040	2236	16320	10/21/2018	0.100	2237	17060	02/14/2019	0.200	2238	17090	02/24/2019	0.300	2239	16410	12/19/2018	0.080 DGS	N/A	34416080A2	02/05/2019	Simulator	Serial Number	Lot Number	Expiration	0.05	SD3967	201603D	03/08/2018	0.08	SD3963	201611B	11/15/2018	0.20	SD3969	201604C	04/05/2018	0.08 DGS	N/A	AG626604	09/22/2018	Department Inspection Performed By _____ ☐ Barometric Pressure _____ Gauge ID# _____ Instrument Mouth Alcohol Solution Lot # _____ Acetone Stock Solution Lot # _____ <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.00</td> <td></td> </tr> <tr> <td>Interferent</td> <td></td> </tr> <tr> <td>0.05</td> <td></td> </tr> <tr> <td>0.08</td> <td></td> </tr> <tr> <td>0.20</td> <td></td> </tr> </tbody> </table> Attachments ☐ Form 41 ☒ Pre-Stability Tests ☐ Flow Calibration ☒ Calibration Adjustment ☒ Post-Stability Tests ☒ Other <u>FDLE/ATP 51</u>	Simulator	Serial Number	0.00		Interferent		0.05		0.08		0.20	
Simulator	Serial Number	Lot Number	Expiration																																																										
0.000	2235	N/A	N/A																																																										
0.040	2236	16320	10/21/2018																																																										
0.100	2237	17060	02/14/2019																																																										
0.200	2238	17090	02/24/2019																																																										
0.300	2239	16410	12/19/2018																																																										
0.080 DGS	N/A	34416080A2	02/05/2019																																																										
Simulator	Serial Number	Lot Number	Expiration																																																										
0.05	SD3967	201603D	03/08/2018																																																										
0.08	SD3963	201611B	11/15/2018																																																										
0.20	SD3969	201604C	04/05/2018																																																										
0.08 DGS	N/A	AG626604	09/22/2018																																																										
Simulator	Serial Number																																																												
0.00																																																													
Interferent																																																													
0.05																																																													
0.08																																																													
0.20																																																													

Notes/Suggested Service: <u>Calibrated to bring the values closer to nominal.</u> <u>0.040 calibration solution failed!</u> <u>Post calibration tests-failed!</u> <u>sg 10/17/17</u>	☐ Instrument Complies with Chapter 11D-8, FAC ☒ Instrument Does Not Comply with Chapter 11D-8, FAC ☐ Return to/Place into Evidentiary Use ☒ Remain Out of Evidentiary Use ☐ Conduct an Agency Inspection Before Evidentiary Use
---	--

Quality Control Review

Date

Revised May 2017

DI-IPS2017R1

FDLE/ATP -048

TYPE OF TEST	SERIAL NUMBER	AGENCY	DATE	PERFORMED BY
Stabilities	80-006924	Broward County Sheriff's Office	8/31/2017	<i>[Signature]</i>

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L
SN: SD3967 Temp: 34.07 0.047 to 0.053 ☒	SN: SD3963 Temp: 34.07c 0.077 to 0.083 ☒	SN: SD3969 Temp: 34.08c 0.194 to 0.206 ☒	Lot AG626604 0.077 to 0.083 ☒
BROWARD COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-006924 08/31/2017 Software: 8100.27 Test g/210L Time Air Blank 0.000 06:12 Control Test 0.047 06:13 Air Blank 0.000 06:13 Control Test 0.047 06:14 Air Blank 0.000 06:15 Control Test 0.048 06:15 Air Blank 0.000 06:16 Control Test Stats Average 0.0473 Std Dev 0.0006 Rel Std Dev(%) 1.2198 <i>[Signature]</i> Operator's Signature	BROWARD COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-006924 08/31/2017 Software: 8100.27 Test g/210L Time Air Blank 0.000 06:17 Control Test 0.077 06:18 Air Blank 0.000 06:19 Control Test 0.078 06:19 Air Blank 0.000 06:20 Control Test 0.077 06:21 Air Blank 0.000 06:21 Control Test Stats Average 0.0773 Std Dev 0.0006 Rel Std Dev(%) 0.7466 <i>[Signature]</i> Operator's Signature	BROWARD COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-006924 08/31/2017 Software: 8100.27 Test g/210L Time Air Blank 0.000 06:22 Control Test 0.193 06:23 Air Blank 0.000 06:24 Control Test 0.195 06:24 Air Blank 0.000 06:25 Control Test 0.196 06:26 Air Blank 0.000 06:26 Control Test Stats Average 0.1947 Std Dev 0.0015 Rel Std Dev(%) 0.7847 <i>[Signature]</i> Operator's Signature	BROWARD COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-006924 08/31/2017 Software: 8100.27 Test g/210L Time Air Blank 0.000 06:28 Control Test 0.082 06:29 Air Blank 0.000 06:29 Control Test 0.082 06:29 Air Blank 0.000 06:30 Control Test 0.082 06:30 Air Blank 0.000 06:31 Control Test Stats Average 0.0820 Std Dev 0.0000 Rel Std Dev(%) 0.0000 <i>[Signature]</i> Operator's Signature

10/17/17
[Signature]

Sol Value = 0.080 g/210L ***
Fit value = 0.3810 mg/l %%%
Samples Taken = 4, Discarded = 1

***** CHANNEL 1

Sample #1 = 3211.00
Sample #2 = 3187.00
Sample #3 = 3169.00
Sample #4 = 3080.00
Average Result = 3145.3333
STD DEV = 57.2916
REL STD DEV = 1.821

***** CHANNEL 2

Sample #1 = 3398.00
Sample #2 = 3393.00
Sample #3 = 3408.00
Sample #4 = 3348.00
Average Result = 3383.0000
STD DEV = 31.2250
REL STD DEV = 0.923

Dry Gas H2O Adjust Results *****

Barometric Pressure = 1015
3 um H2O Adjust (mg/l*10,000) = 664
9 um H2O Adjust (mg/l*10,000) = 426
**** AUTO CAL PASS

Optical Calibration Continued	
SN: 80-006924	
Agency: Broward County SO	
Date: 8/31/2017	
Quadratic Fit: +/-0.002g/210L	
By: <i>ALL</i>	



Auto calibration indicated failed and requested to re-run one of the solutions.

10/17/17
JG

TYPE OF TEST	SERIAL NUMBER	AGENCY	DATE	PERFORMED BY
Post Stabilities	80-006924	Broward County Sheriff's Office	8/31/2017	<i>Paul</i>

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L																																																																																																																																																
SN: SD3967 Temp: 34.07 0.047 to 0.053 ☒	SN: SD3963 Temp: 34.07 0.077 to 0.083 ☒	SN: SD3969 Temp: 34.08 0.194 to 0.206 ☒	AG26604 0.077 to 0.083 ☒																																																																																																																																																
BROWARD COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 08/31/2017 Software: 8100.27 <table><tr><th>Test</th><th>g/210L</th><th>Time</th></tr><tr><td>Air Blank</td><td>0.000</td><td>08:24</td></tr><tr><td>Control Test</td><td>0.050</td><td>08:24</td></tr><tr><td>Air Blank</td><td>0.000</td><td>08:25</td></tr><tr><td>Control Test</td><td>0.051</td><td>08:25</td></tr><tr><td>Air Blank</td><td>0.000</td><td>08:26</td></tr><tr><td>Control Test</td><td>0.050</td><td>08:27</td></tr><tr><td>Air Blank</td><td>0.000</td><td>08:27</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0503</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>1.1471</td><td></td></tr></table>	Test	g/210L	Time	Air Blank	0.000	08:24	Control Test	0.050	08:24	Air Blank	0.000	08:25	Control Test	0.051	08:25	Air Blank	0.000	08:26	Control Test	0.050	08:27	Air Blank	0.000	08:27	Control Test Stats			Average	0.0503		Std Dev	0.0006		Rel Std Dev(%)	1.1471		BROWARD COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 08/31/2017 Software: 8100.27 <table><tr><th>Test</th><th>g/210L</th><th>Time</th></tr><tr><td>Air Blank</td><td>0.000</td><td>08:28</td></tr><tr><td>Control Test</td><td>0.082</td><td>08:29</td></tr><tr><td>Air Blank</td><td>0.000</td><td>08:29</td></tr><tr><td>Control Test</td><td>0.082</td><td>08:30</td></tr><tr><td>Air Blank</td><td>0.000</td><td>08:31</td></tr><tr><td>Control Test</td><td>0.084</td><td>08:31</td></tr><tr><td>Air Blank</td><td>0.000</td><td>08:32</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0827</td><td></td></tr><tr><td>Std Dev</td><td>0.0012</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>1.3968</td><td></td></tr></table>	Test	g/210L	Time	Air Blank	0.000	08:28	Control Test	0.082	08:29	Air Blank	0.000	08:29	Control Test	0.082	08:30	Air Blank	0.000	08:31	Control Test	0.084	08:31	Air Blank	0.000	08:32	Control Test Stats			Average	0.0827		Std Dev	0.0012		Rel Std Dev(%)	1.3968		BROWARD COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 08/31/2017 Software: 8100.27 <table><tr><th>Test</th><th>g/210L</th><th>Time</th></tr><tr><td>Air Blank</td><td>0.000</td><td>08:38</td></tr><tr><td>Control Test</td><td>0.201</td><td>08:38</td></tr><tr><td>Air Blank</td><td>0.000</td><td>08:39</td></tr><tr><td>Control Test</td><td>0.200</td><td>08:40</td></tr><tr><td>Air Blank</td><td>0.000</td><td>08:40</td></tr><tr><td>Control Test</td><td>0.201</td><td>08:41</td></tr><tr><td>Air Blank</td><td>0.000</td><td>08:41</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.2007</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.2877</td><td></td></tr></table>	Test	g/210L	Time	Air Blank	0.000	08:38	Control Test	0.201	08:38	Air Blank	0.000	08:39	Control Test	0.200	08:40	Air Blank	0.000	08:40	Control Test	0.201	08:41	Air Blank	0.000	08:41	Control Test Stats			Average	0.2007		Std Dev	0.0006		Rel Std Dev(%)	0.2877		BROWARD COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 08/31/2017 Software: 8100.27 <table><tr><th>Test</th><th>g/210L</th><th>Time</th></tr><tr><td>Air Blank</td><td>0.000</td><td>08:44</td></tr><tr><td>Control Test</td><td>0.079</td><td>08:44</td></tr><tr><td>Air Blank</td><td>0.000</td><td>08:45</td></tr><tr><td>Control Test</td><td>0.079</td><td>08:45</td></tr><tr><td>Air Blank</td><td>0.000</td><td>08:46</td></tr><tr><td>Control Test</td><td>0.079</td><td>08:46</td></tr><tr><td>Air Blank</td><td>0.000</td><td>08:47</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0790</td><td></td></tr><tr><td>Std Dev</td><td>0.0000</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr></table>	Test	g/210L	Time	Air Blank	0.000	08:44	Control Test	0.079	08:44	Air Blank	0.000	08:45	Control Test	0.079	08:45	Air Blank	0.000	08:46	Control Test	0.079	08:46	Air Blank	0.000	08:47	Control Test Stats			Average	0.0790		Std Dev	0.0000		Rel Std Dev(%)	0.0000	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	08:24																																																																																																																																																	
Control Test	0.050	08:24																																																																																																																																																	
Air Blank	0.000	08:25																																																																																																																																																	
Control Test	0.051	08:25																																																																																																																																																	
Air Blank	0.000	08:26																																																																																																																																																	
Control Test	0.050	08:27																																																																																																																																																	
Air Blank	0.000	08:27																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0503																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	1.1471																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	08:28																																																																																																																																																	
Control Test	0.082	08:29																																																																																																																																																	
Air Blank	0.000	08:29																																																																																																																																																	
Control Test	0.082	08:30																																																																																																																																																	
Air Blank	0.000	08:31																																																																																																																																																	
Control Test	0.084	08:31																																																																																																																																																	
Air Blank	0.000	08:32																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0827																																																																																																																																																		
Std Dev	0.0012																																																																																																																																																		
Rel Std Dev(%)	1.3968																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	08:38																																																																																																																																																	
Control Test	0.201	08:38																																																																																																																																																	
Air Blank	0.000	08:39																																																																																																																																																	
Control Test	0.200	08:40																																																																																																																																																	
Air Blank	0.000	08:40																																																																																																																																																	
Control Test	0.201	08:41																																																																																																																																																	
Air Blank	0.000	08:41																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.2007																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.2877																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	08:44																																																																																																																																																	
Control Test	0.079	08:44																																																																																																																																																	
Air Blank	0.000	08:45																																																																																																																																																	
Control Test	0.079	08:45																																																																																																																																																	
Air Blank	0.000	08:46																																																																																																																																																	
Control Test	0.079	08:46																																																																																																																																																	
Air Blank	0.000	08:47																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0790																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
<i>Paul</i> Operator's Signature 10/17/17	<i>Paul</i> Operator's Signature	<i>Paul</i> Operator's Signature	<i>Paul</i> Operator's Signature																																																																																																																																																

TYPE OF TEST	SERIAL NUMBER	AGENCY	DATE	PERFORMED BY
Post Stabilities 2 nd Test	80-006924	Broward County Sheriff's Office	8/31/2017	<i>mu</i>

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L																																				
0.047 to 0.053 ☐	0.077 to 0.083 ☒	0.194 to 0.206 ☐	0.077 to 0.083 ☐																																				
	BROWARD COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 08/31/2017 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>08:33</td></tr> <tr><td>Control Test</td><td>0.084</td><td>08:34</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>08:34</td></tr> <tr><td>Control Test</td><td>0.082</td><td>08:35</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>08:36</td></tr> <tr><td>Control Test</td><td>0.083</td><td>08:36</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>08:37</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0830</td><td></td></tr> <tr><td>Std Dev</td><td>0.0010</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>1.2048</td><td></td></tr> </tbody> </table> <i>mu</i> Operator's Signature	Test	g/210L	Time	Air Blank	0.000	08:33	Control Test	0.084	08:34	Air Blank	0.000	08:34	Control Test	0.082	08:35	Air Blank	0.000	08:36	Control Test	0.083	08:36	Air Blank	0.000	08:37	Control Test Stats			Average	0.0830		Std Dev	0.0010		Rel Std Dev(%)	1.2048			
Test	g/210L	Time																																					
Air Blank	0.000	08:33																																					
Control Test	0.084	08:34																																					
Air Blank	0.000	08:34																																					
Control Test	0.082	08:35																																					
Air Blank	0.000	08:36																																					
Control Test	0.083	08:36																																					
Air Blank	0.000	08:37																																					
Control Test Stats																																							
Average	0.0830																																						
Std Dev	0.0010																																						
Rel Std Dev(%)	1.2048																																						

10/17/17
ce