Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: DEVAUL,ERIC
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 01/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1870563
DRY GAS EXPIRY DATE: 06/17/2018
VIOLATION CODE: DUI
ARREST TIME: 01:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:21

SUBJECT'S LAST NAME: ROWN

DATE/TIME OF TEST: 01/30/2016 02:59:40

Diagnostics Check		OK	02:53
Air Blank		0.000	02:53
Control Test		0.080	02:53
Air Blank		0.000	02:54
Subject Sample	# 1	0.166	02:55
Breath Volume		2.570	02:55
Air Blank		0.000	02:55
Air Blank		0.000	02:57
Subject Sample	# 2	0.164	02:58
Breath Volume		2.281	02:58
Air Blank		0.000	02:58
Control Test		0.080	02:59
Air Blank		0.000	02:59
Diagnostics Check		OK	02:59

Data Download Date/Time: 02/15/2016 11:54

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAKAI,MICHAEL,T
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: MAKAI,MICHAEL
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1841081
DRY GAS EXPIRY DATE: 04/08/2018
VIOLATION CODE: DUI
ARREST TIME: 00:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:52

SUBJECT'S LAST NAME: MAYS

DATE/TIME OF TEST: 01/01/2016 01:19:03

Diagnostics Check		OK	01:16	
Air Blank		0.000	01:16	
Control Test		0.080	01:16	
Air Blank		0.000	01:17	
Subject Sample	# 1		01:17	Subject Test Refused
Breath Volume		0.000	01:17	
Air Blank		0.000	01:18	
Control Test		0.080	01:18	
Air Blank		0.000	01:18	
Diagnostics Check		OK	01:19	

Data Download Date/Time: 01/25/2016 10:00

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TATE,ROBERT,J
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: TATE,ROBERT
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1841081
DRY GAS EXPIRY DATE: 04/08/2018
VIOLATION CODE: DUI
ARREST TIME: 02:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: DAVENPORT

DATE/TIME OF TEST: 01/09/2016 04:18:22

Diagnostics Check		OK	04:15	
Air Blank		0.000	04:15	
Control Test		0.080	04:16	
Air Blank		0.000	04:16	
Subject Sample	# 1		04:16	Subject Test Refused
Breath Volume		0.000	04:16	
Air Blank		0.000	04:17	
Control Test		0.080	04:17	
Air Blank		0.000	04:18	
Diagnostics Check		OK	04:18	

Data Download Date/Time: 01/25/2016 10:00

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DE LIMA,ALEX,
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: DE LIMA,ALEX
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1841081
DRY GAS EXPIRY DATE: 04/08/2018
VIOLATION CODE: DUI
ARREST TIME: 14:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:10

SUBJECT'S LAST NAME: MCMURRAY

DATE/TIME OF TEST: 01/10/2016 15:37:13

Diagnostics Check		OK	15:30
Air Blank		0.000	15:31
Control Test		0.080	15:31
Air Blank		0.000	15:32
Subject Sample	# 1	0.210	15:32
Breath Volume		1.910	15:32
Air Blank		0.000	15:33
Air Blank		0.000	15:35
Subject Sample	# 2	0.195	15:35
Breath Volume		1.917	15:35
Air Blank		0.000	15:36
Control Test		0.079	15:36
Air Blank		0.000	15:37
Diagnostics Check		OK	15:37

Data Download Date/Time: 01/25/2016 10:00

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TATE,ROBERT,J
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: TATE,ROBERT
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1841081
DRY GAS EXPIRY DATE: 04/08/2018
VIOLATION CODE: DUI
ARREST TIME: 22:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: VALDOVINOS AGUIRRE

DATE/TIME OF TEST: 01/11/2016 00:23:25

Diagnostics Check		OK	00:20
Air Blank		0.000	00:20
Control Test		0.081	00:21
Air Blank		0.000	00:21
Subject Sample	# 1	0.013	00:22
Breath Volume		1.714	00:22
Air Blank		0.000	00:22
Air Blank		0.000	00:23

Data Download Date/Time: 01/25/2016 10:00

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAKAI,MICHAEL,T
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: MAKAI,MICHAEL
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1841081
DRY GAS EXPIRY DATE: 04/08/2018
VIOLATION CODE: DUI
ARREST TIME: 23:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: MACINTYRE

DATE/TIME OF TEST: 01/14/2016 01:03:58

Diagnostics Check		OK	01:00	
Air Blank		0.000	01:01	
Control Test		0.081	01:01	
Air Blank		0.000	01:01	
Subject Sample	# 1		01:02	Subject Test Refused
Breath Volume		0.000	01:02	
Air Blank		0.000	01:03	
Control Test		0.081	01:03	
Air Blank		0.000	01:03	
Diagnostics Check		OK	01:03	

Data Download Date/Time: 01/25/2016 10:00

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLE,DENNIS,A
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: COLE,DENNIS
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1841081
DRY GAS EXPIRY DATE: 04/08/2018
VIOLATION CODE: DUI
ARREST TIME: 04:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:40

SUBJECT'S LAST NAME: REEVES

DATE/TIME OF TEST: 01/16/2016 05:12:08

Diagnostics Check		OK	05:04
Air Blank		0.000	05:05
Control Test		0.080	05:05
Air Blank		0.000	05:06
Subject Sample	# 1	0.169	05:07
Breath Volume		1.566	05:07
Air Blank		0.000	05:08
Air Blank		0.000	05:09
Subject Sample	# 2	0.168	05:10
Breath Volume		1.140	05:10
Air Blank		0.000	05:11
Control Test		0.078	05:11
Air Blank		0.000	05:12
Diagnostics Check		OK	05:12

Data Download Date/Time: 01/25/2016 10:00

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: SANTANA,ELIER
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1841081
DRY GAS EXPIRY DATE: 04/08/2018
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: ARITA

DATE/TIME OF TEST: 01/17/2016 04:42:39

Diagnostics Check		OK	04:32
Air Blank		0.000	04:32
Control Test		0.080	04:33
Air Blank		0.000	04:33
Subject Sample	# 1	0.250	04:34
Breath Volume		2.175	04:34
Air Blank		0.000	04:35
Air Blank		0.000	04:37
Subject Sample	# 2	0.273	04:37
Breath Volume		2.042	04:37
Air Blank		0.000	04:38
Air Blank		0.000	04:40
Subject Sample	# 3	0.278	04:41
Breath Volume		2.183	04:41
Air Blank		0.000	04:41
Control Test		0.077	04:42
Air Blank		0.000	04:42
Diagnostics Check		OK	04:42

Data Download Date/Time: 01/25/2016 10:00

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001865

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTLE,CORY,A
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: SUTTLE,CORY
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 01/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 27713080A1
DRY GAS EXPIRY DATE: 11/01/2015
VIOLATION CODE: DUI
ARREST TIME: 02:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: BURDGE

DATE/TIME OF TEST: 01/28/2016 04:05:18

Diagnostics Check		OK	03:56
Air Blank		0.000	03:57
Control Test		0.079	03:57
Air Blank		0.000	03:57
Subject Sample	# 1	0.162	04:00
Breath Volume		2.398	04:00
Air Blank		0.000	04:00
Air Blank		0.000	04:02
Subject Sample	# 2	0.174	04:03
Breath Volume		2.066	04:03
Air Blank		0.000	04:04
Control Test		0.080	04:04
Air Blank		0.000	04:05
Diagnostics Check		OK	04:05

Data Download Date/Time: 02/15/2016 09:48

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002231

INSTRUMENT REGISTERED TO: NEW SMYRNA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAGE,BRIAN,R
BREATH TEST OPERATOR AGENCY: New Smyrna Beach PD
ARRESTING OFFICER: PAGE,BRIAN
ARRESTING OFFICER AGENCY: New Smyrna Beach PD
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12415080A5
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:44

SUBJECT'S LAST NAME: KLISH

DATE/TIME OF TEST: 01/07/2016 22:10:35

Diagnostics Check		OK	22:04
Air Blank		0.000	22:04
Control Test		0.078	22:05
Air Blank		0.000	22:05
Subject Sample	# 1	0.212	22:06
Breath Volume		1.804	22:06
Air Blank		0.000	22:06
Air Blank		0.000	22:08
Subject Sample	# 2	0.222	22:09
Breath Volume		1.625	22:09
Air Blank		0.000	22:09
Control Test		0.078	22:09
Air Blank		0.000	22:10
Diagnostics Check		OK	22:10

Data Download Date/Time: 01/30/2016 09:16

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002231

INSTRUMENT REGISTERED TO: NEW SMYRNA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRKLAND,RICHARD,E
BREATH TEST OPERATOR AGENCY: New Smyrna Beach PD
ARRESTING OFFICER: KIRKLAND,RICHARD
ARRESTING OFFICER AGENCY: New Smyrna Beach PD
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12415080A5
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:54

SUBJECT'S LAST NAME: GIBSON

DATE/TIME OF TEST: 01/14/2016 22:24:05

Diagnostics Check		OK	22:14
Air Blank		0.000	22:15
Control Test		0.078	22:15
Air Blank		0.000	22:16
Subject Sample	# 1	0.192	22:16
Breath Volume		1.425	22:16
Air Blank		0.000	22:17
Air Blank		0.000	22:19
Subject Sample	# 2	0.193	22:22
Breath Volume		1.574	22:22
Air Blank		0.000	22:23
Control Test		0.078	22:23
Air Blank		0.000	22:24
Diagnostics Check		OK	22:24

Data Download Date/Time: 01/30/2016 09:16

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002231

INSTRUMENT REGISTERED TO: NEW SMYRNA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOYCE,VALERIE,J
BREATH TEST OPERATOR AGENCY: New Smyrna Beach PD
ARRESTING OFFICER: DALY,BEVERLY
ARRESTING OFFICER AGENCY: New Smyrna Beach PD
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12415080A5
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 07:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:14

SUBJECT'S LAST NAME: OSHAUGHNESSY

DATE/TIME OF TEST: 01/17/2016 08:42:03

Diagnostics Check		OK	08:35
Air Blank		0.000	08:35
Control Test		0.077	08:35
Air Blank		0.000	08:36
Subject Sample	# 1	0.231	08:37
Breath Volume		1.488	08:37
Air Blank		0.000	08:37
Air Blank		0.000	08:39
Subject Sample	# 2	0.217	08:40
Breath Volume		1.898	08:40
Air Blank		0.000	08:41
Control Test		0.077	08:41
Air Blank		0.000	08:41
Diagnostics Check		OK	08:42

Data Download Date/Time: 01/30/2016 09:16

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002231

INSTRUMENT REGISTERED TO: NEW SMYRNA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRKLAND,RICHARD,E
BREATH TEST OPERATOR AGENCY: New Smyrna Beach PD
ARRESTING OFFICER: JENKINS,ANDRE
ARRESTING OFFICER AGENCY: New Smyrna Beach PD
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12415080A5
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: PAREKH

DATE/TIME OF TEST: 01/24/2016 03:20:49

Diagnostics Check		OK	03:11
Air Blank		0.000	03:11
Control Test		0.079	03:11
Air Blank		0.000	03:12
Subject Sample	# 1	0.226	03:14
Breath Volume		1.296	03:14
Air Blank		0.000	03:15
Air Blank		0.000	03:17
Subject Sample	# 2	0.208	03:19
Breath Volume		2.121	03:19
Air Blank		0.000	03:19
Control Test		0.078	03:20
Air Blank		0.000	03:20
Diagnostics Check		OK	03:20

Data Download Date/Time: 01/30/2016 09:16

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002231

INSTRUMENT REGISTERED TO: NEW SMYRNA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SEVEANCE,MARK,D
BREATH TEST OPERATOR AGENCY: New Smyrna Beach PD
ARRESTING OFFICER: HUPERT,CHUCK
ARRESTING OFFICER AGENCY: New Smyrna Beach PD
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12415080A5
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 13:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:57

SUBJECT'S LAST NAME: EDWARDS-POLAK

DATE/TIME OF TEST: 01/27/2016 14:46:07

Diagnostics Check		OK	14:39
Air Blank		0.000	14:40
Control Test		0.078	14:40
Air Blank		0.000	14:41
Subject Sample	# 1	0.000	14:41
Breath Volume		1.871	14:41
Air Blank		0.000	14:42
Air Blank		0.000	14:44
Subject Sample	# 2	0.000	14:44
Breath Volume		2.355	14:44
Air Blank		0.000	14:45
Control Test		0.077	14:45
Air Blank		0.000	14:46
Diagnostics Check		OK	14:46

Data Download Date/Time: 01/30/2016 09:16

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002232

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERNANDEZ,JAVIER,
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: FERNANDEZ,JAVIER
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 12/14/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482986
DRY GAS EXPIRY DATE: 04/03/2016
VIOLATION CODE: DUI
ARREST TIME: 15:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:15

SUBJECT'S LAST NAME: MONTOYA

DATE/TIME OF TEST: 01/01/2016 15:44:09

Diagnostics Check		OK	15:35
Air Blank		0.000	15:36
Control Test		0.081	15:36
Air Blank		0.000	15:37
Subject Sample	# 1	0.167	15:39
Breath Volume		1.855	15:39
Air Blank		0.000	15:40
Air Blank		0.000	15:41
Subject Sample	# 2	0.171	15:42
Breath Volume		1.914	15:42
Air Blank		0.000	15:43
Control Test		0.079	15:43
Air Blank		0.000	15:44
Diagnostics Check		OK	15:44

Data Download Date/Time: 01/26/2016 10:15

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002232

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CREWS,CURTIS,L
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: DELGADO,JANELLE
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 12/14/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482986
DRY GAS EXPIRY DATE: 04/03/2016
VIOLATION CODE: DUI
ARREST TIME: 03:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: ROMERO

DATE/TIME OF TEST: 01/04/2016 03:59:27

Diagnostics Check		OK	03:56	
Air Blank		0.000	03:57	
Control Test		0.080	03:57	
Air Blank		0.000	03:57	
Subject Sample	# 1		03:58	Subject Test Refused
Breath Volume		0.000	03:58	
Air Blank		0.000	03:58	
Control Test		0.081	03:58	
Air Blank		0.000	03:59	
Diagnostics Check		OK	03:59	

Data Download Date/Time: 01/26/2016 10:15

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002232

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAYEGH,MICHELLE,M
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: CARPIO,MELANIE
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 12/14/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482986
DRY GAS EXPIRY DATE: 04/03/2016
VIOLATION CODE: DUI
ARREST TIME: 06:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:58

SUBJECT'S LAST NAME: ALVAREZ

DATE/TIME OF TEST: 01/10/2016 07:30:56

Diagnostics Check		OK	07:24
Air Blank		0.000	07:24
Control Test		0.081	07:25
Air Blank		0.000	07:25
Subject Sample	# 1	0.083	07:26
Breath Volume		1.347	07:26
Air Blank		0.000	07:27
Air Blank		0.000	07:28
Subject Sample	# 2	0.083	07:29
Breath Volume		2.503	07:29
Air Blank		0.000	07:30
Control Test		0.080	07:30
Air Blank		0.000	07:30
Diagnostics Check		OK	07:30

Data Download Date/Time: 01/26/2016 10:15

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002232

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUIJANO,JEANETTE,
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: SUAREZ,RONALD
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 12/14/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482986
DRY GAS EXPIRY DATE: 04/03/2016
VIOLATION CODE: DUI
ARREST TIME: 14:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:50

SUBJECT'S LAST NAME: MAYO

DATE/TIME OF TEST: 01/10/2016 15:18:27

Diagnostics Check		OK	15:11
Air Blank		0.000	15:11
Control Test		0.080	15:12
Air Blank		0.000	15:12
Subject Sample	# 1	0.288	15:13
Breath Volume		1.878	15:13
Air Blank		0.000	15:14
Air Blank		0.000	15:16
Subject Sample	# 2	0.300	15:16
Breath Volume		3.347	15:16
Air Blank		0.000	15:17
Control Test		0.078	15:17
Air Blank		0.000	15:18
Diagnostics Check		OK	15:18

Data Download Date/Time: 01/26/2016 10:15

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002232

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,JOSE,D
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: VILLA,LAURENCE
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 12/14/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482986
DRY GAS EXPIRY DATE: 04/03/2016
VIOLATION CODE: DUI
ARREST TIME: 00:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: CARRERASMARIN

DATE/TIME OF TEST: 01/11/2016 01:36:27

Diagnostics Check		OK	01:29
Air Blank		0.000	01:30
Control Test		0.081	01:30
Air Blank		0.000	01:31
Subject Sample	# 1	0.187	01:31
Breath Volume		1.226	01:31
Air Blank		0.000	01:32
Air Blank		0.000	01:34
Subject Sample	# 2	0.171	01:34
Breath Volume		2.957	01:34
Air Blank		0.000	01:35
Control Test		0.079	01:35
Air Blank		0.000	01:36
Diagnostics Check		OK	01:36

Data Download Date/Time: 01/26/2016 10:15

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002232

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CREWS,CURTIS,L
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: PEREZ,JOSE
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 12/14/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482986
DRY GAS EXPIRY DATE: 04/03/2016
VIOLATION CODE: DUI
ARREST TIME: 06:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:45

SUBJECT'S LAST NAME: WELLS

DATE/TIME OF TEST: 01/11/2016 07:14:09

Diagnostics Check		OK	07:10	
Air Blank		0.000	07:10	
Control Test		0.080	07:11	
Air Blank		0.000	07:11	
Subject Sample	# 1		07:12	Subject Test Refused
Breath Volume		0.000	07:12	
Air Blank		0.000	07:13	
Control Test		0.080	07:13	
Air Blank		0.000	07:14	
Diagnostics Check		OK	07:14	

Data Download Date/Time: 01/26/2016 10:15

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002232

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUZMAN,ANTHONY,
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: POMARES,MAKIN
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 12/14/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482986
DRY GAS EXPIRY DATE: 04/03/2016
VIOLATION CODE: DUI
ARREST TIME: 11:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:32

SUBJECT'S LAST NAME: SUAREZ

DATE/TIME OF TEST: 01/15/2016 13:34:35

Diagnostics Check		OK	13:31	
Air Blank		0.000	13:32	
Control Test		0.080	13:32	
Air Blank		0.000	13:33	
Subject Sample	# 1		13:33	Subject Test Refused
Breath Volume		0.000	13:33	
Air Blank		0.000	13:33	
Control Test		0.081	13:34	
Air Blank		0.000	13:34	
Diagnostics Check		OK	13:34	

Data Download Date/Time: 01/26/2016 10:15

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002232

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUZMAN,ANTHONY,
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: DOMINGUEZ,ERIC
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 12/14/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482986
DRY GAS EXPIRY DATE: 04/03/2016
VIOLATION CODE: DUI
ARREST TIME: 23:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:14

SUBJECT'S LAST NAME: ALIYA

DATE/TIME OF TEST: 01/22/2016 23:46:53

Diagnostics Check		OK	23:39
Air Blank		0.000	23:39
Control Test		0.080	23:40
Air Blank		0.000	23:40
Subject Sample	# 1	0.131	23:42
Breath Volume		1.593	23:42
Air Blank		0.000	23:43
Air Blank		0.000	23:45
Subject Sample	# 2	0.131	23:45
Breath Volume		1.425	23:45
Air Blank		0.000	23:46
Control Test		0.081	23:46
Air Blank		0.000	23:46
Diagnostics Check		OK	23:46

Data Download Date/Time: 01/26/2016 10:15

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002232

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,RICHARD,
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: RODRIGUEZ,RICHARD
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 12/14/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482986
DRY GAS EXPIRY DATE: 04/03/2016
VIOLATION CODE: DUI
ARREST TIME: 01:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:18

SUBJECT'S LAST NAME: VELASQUEZ

DATE/TIME OF TEST: 01/23/2016 01:42:38

Diagnostics Check		OK	01:39	
Air Blank		0.000	01:40	
Control Test		0.079	01:40	
Air Blank		0.000	01:40	
Subject Sample	# 1		01:41	Subject Test Refused
Breath Volume		0.000	01:41	
Air Blank		0.000	01:41	
Control Test		0.080	01:42	
Air Blank		0.000	01:42	
Diagnostics Check		OK	01:42	

Data Download Date/Time: 01/26/2016 10:15

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002232

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERNANDEZ,JAVIER,
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: GUASTO,NICHOLAS
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 12/14/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482986
DRY GAS EXPIRY DATE: 04/03/2016
VIOLATION CODE: DUI
ARREST TIME: 03:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: GUTIERREZ

DATE/TIME OF TEST: 01/23/2016 04:03:42

Diagnostics Check		OK	03:56
Air Blank		0.000	03:56
Control Test		0.081	03:56
Air Blank		0.000	03:57
Subject Sample	# 1	0.110	03:58
Breath Volume		1.246	03:58
Air Blank		0.000	03:59
Air Blank		0.000	04:01
Subject Sample	# 2	0.107	04:02
Breath Volume		1.234	04:02
Air Blank		0.000	04:02
Control Test		0.079	04:03
Air Blank		0.000	04:03
Diagnostics Check		OK	04:03

Data Download Date/Time: 01/26/2016 10:15

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002232

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERNANDEZ,JAVIER,
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: FERNANDEZ,JAVIER
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 12/14/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482986
DRY GAS EXPIRY DATE: 04/03/2016
VIOLATION CODE: DUI
ARREST TIME: 21:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:27

SUBJECT'S LAST NAME: PELUSO

DATE/TIME OF TEST: 01/24/2016 22:07:10

Diagnostics Check		OK	21:59
Air Blank		0.000	21:59
Control Test		0.081	22:00
Air Blank		0.000	22:00
Subject Sample	# 1	0.091	22:02
Breath Volume		1.542	22:02
Air Blank		0.000	22:03
Air Blank		0.000	22:05
Subject Sample	# 2	0.093	22:05
Breath Volume		1.585	22:05
Air Blank		0.000	22:06
Control Test		0.081	22:06
Air Blank		0.000	22:07
Diagnostics Check		OK	22:07

Data Download Date/Time: 01/26/2016 10:15

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002232

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,JOSE,D
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: COLOMBO,DANIEL
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 12/14/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482986
DRY GAS EXPIRY DATE: 04/03/2016
VIOLATION CODE: DUI
ARREST TIME: 23:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: EDWARDS

DATE/TIME OF TEST: 01/25/2016 00:44:00

Diagnostics Check		OK	00:37
Air Blank		0.000	00:38
Control Test		0.081	00:38
Air Blank		0.000	00:39
Subject Sample	# 1	0.191	00:39
Breath Volume		2.160	00:39
Air Blank		0.000	00:40
Air Blank		0.000	00:42
Subject Sample	# 2	0.185	00:42
Breath Volume		2.359	00:42
Air Blank		0.000	00:43
Control Test		0.080	00:43
Air Blank		0.000	00:43
Diagnostics Check		OK	00:44

Data Download Date/Time: 01/26/2016 10:15

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002232

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUASTO,NICHOLAS,L
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: MILLAN,KEVIN
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 01/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482986
DRY GAS EXPIRY DATE: 04/03/2016
VIOLATION CODE: DUI
ARREST TIME: 02:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:22

SUBJECT'S LAST NAME: RADUENZ

DATE/TIME OF TEST: 01/31/2016 02:55:53

Diagnostics Check		OK	02:48
Air Blank		0.000	02:48
Control Test		0.081	02:49
Air Blank		0.000	02:49
Subject Sample	# 1	0.136	02:50
Breath Volume		1.960	02:50
Air Blank		0.000	02:51
Air Blank		0.000	02:53
Subject Sample	# 2	0.133	02:54
Breath Volume		1.886	02:54
Air Blank		0.000	02:54
Control Test		0.081	02:55
Air Blank		0.000	02:55
Diagnostics Check		OK	02:55

Data Download Date/Time: 02/23/2016 10:10

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002232

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SOCARRAS,PEDRO,A
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: DOMINGUEZ,ERIC
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 01/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482986
DRY GAS EXPIRY DATE: 04/03/2016
VIOLATION CODE: DUI
ARREST TIME: 03:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:54

SUBJECT'S LAST NAME: JANJUA

DATE/TIME OF TEST: 01/31/2016 04:20:48

Diagnostics Check		OK	04:14
Air Blank		0.000	04:15
Control Test		0.080	04:15
Air Blank		0.000	04:15
Subject Sample	# 1	0.094	04:16
Breath Volume		1.308	04:16
Air Blank		0.000	04:16
Air Blank		0.000	04:18
Subject Sample	# 2	0.097	04:19
Breath Volume		1.429	04:19
Air Blank		0.000	04:19
Control Test		0.080	04:20
Air Blank		0.000	04:20
Diagnostics Check		OK	04:20

Data Download Date/Time: 02/23/2016 10:10

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002232

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLAN,KEVIN,
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: GUASTO,NICHOLAS
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 01/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482986
DRY GAS EXPIRY DATE: 04/03/2016
VIOLATION CODE: DUI
ARREST TIME: 04:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:10

SUBJECT'S LAST NAME: SCHLAMAN

DATE/TIME OF TEST: 01/31/2016 05:41:50

Diagnostics Check		OK	05:33
Air Blank		0.000	05:33
Control Test		0.080	05:34
Air Blank		0.000	05:34
Subject Sample	# 1	0.180	05:36
Breath Volume		1.523	05:36
Air Blank		0.000	05:37
Air Blank		0.000	05:39
Subject Sample	# 2	0.164	05:40
Breath Volume		1.214	05:40
Air Blank		0.000	05:40
Control Test		0.080	05:41
Air Blank		0.000	05:41
Diagnostics Check		OK	05:41

Data Download Date/Time: 02/23/2016 10:10

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002463

INSTRUMENT REGISTERED TO: HIALEAH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHAVEZ,MAYRELIS,
BREATH TEST OPERATOR AGENCY: Hialeah PD
ARRESTING OFFICER: FARKAS,ABRAHAM
ARRESTING OFFICER AGENCY: Hialeah PD
LAST AGENCY INSPECTION: 12/10/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06415080A1
DRY GAS EXPIRY DATE: 04/05/2017
VIOLATION CODE: DUI
ARREST TIME: 11:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:30

SUBJECT'S LAST NAME: ALFONSO

DATE/TIME OF TEST: 01/02/2016 12:56:52

Diagnostics Check		OK	12:51
Air Blank		0.000	12:51
Control Test		0.075	12:51
Air Blank		0.000	12:52
Subject Sample	# 1	0.000	12:52
Breath Volume		1.898	12:52
Air Blank		0.000	12:53
Air Blank		0.000	12:55
Subject Sample	# 2	0.000	12:55
Breath Volume		1.738	12:55
Air Blank		0.000	12:55
Control Test		0.076	12:56
Air Blank		0.000	12:56
Diagnostics Check		OK	12:56

Data Download Date/Time: 01/21/2016 11:03

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002463

INSTRUMENT REGISTERED TO: HIALEAH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JANVIER,FRITZ,G
BREATH TEST OPERATOR AGENCY: Hialeah PD
ARRESTING OFFICER: MORALES,M
ARRESTING OFFICER AGENCY: Hialeah PD
LAST AGENCY INSPECTION: 12/10/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06415080A1
DRY GAS EXPIRY DATE: 04/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: CARMONA VARGAS

DATE/TIME OF TEST: 01/19/2016 23:14:50

Diagnostics Check		OK	23:10	
Air Blank		0.000	23:10	
Control Test		0.080	23:10	
Air Blank		0.000	23:11	
Subject Sample	# 1		23:13	Subject Test Refused
Breath Volume		0.000	23:13	
Air Blank		0.000	23:13	
Control Test		0.079	23:14	
Air Blank		0.000	23:14	
Diagnostics Check		OK	23:14	

Data Download Date/Time: 01/21/2016 11:03

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002464

INSTRUMENT REGISTERED TO: HIALEAH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JANVIER,FRITZ,G
BREATH TEST OPERATOR AGENCY: Hialeah PD
ARRESTING OFFICER: JANVIER,FRITZ
ARRESTING OFFICER AGENCY: Hialeah PD
LAST AGENCY INSPECTION: 12/10/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06415080A1
DRY GAS EXPIRY DATE: 04/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:35

SUBJECT'S LAST NAME: FINALES CRUZ

DATE/TIME OF TEST: 01/08/2016 22:02:45

Diagnostics Check		OK	21:55
Air Blank		0.000	21:55
Control Test		0.081	21:56
Air Blank		0.000	21:56
Subject Sample	# 1	0.066	21:58
Breath Volume		1.265	21:58
Air Blank		0.000	21:58
Air Blank		0.000	22:00
Subject Sample	# 2	0.065	22:01
Breath Volume		1.109	22:01
Air Blank		0.000	22:01
Control Test		0.078	22:02
Air Blank		0.000	22:02
Diagnostics Check		OK	22:02

Data Download Date/Time: 01/21/2016 11:16

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002987

INSTRUMENT REGISTERED TO: MIAMI GARDENS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAMAYO,KEVIN,X
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: TAMAYO,KEVIN
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 12/02/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A4
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: BERDUGO

DATE/TIME OF TEST: 01/12/2016 01:24:50

Diagnostics Check	OK	01:24	
Air Blank		01:24	RFI Detect
Air Blank	0.000	01:24	

Data Download Date/Time: 02/02/2016 08:07

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002987

INSTRUMENT REGISTERED TO: MIAMI GARDENS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAMAYO,KEVIN,X
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: TAMAYO,KEVIN
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 12/02/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A4
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: BERDUGO

DATE/TIME OF TEST: 01/12/2016 01:30:27

Diagnostics Check		OK	01:27	
Air Blank		0.000	01:27	
Control Test		0.076	01:28	
Air Blank		0.000	01:28	
Subject Sample	# 1		01:29	Subject Test Refused
Breath Volume		0.000	01:29	
Air Blank		0.000	01:29	
Control Test		0.077	01:29	
Air Blank		0.000	01:30	
Diagnostics Check		OK	01:30	

Data Download Date/Time: 02/02/2016 08:07

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002988

INSTRUMENT REGISTERED TO: MIAMI GARDENS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADDISON,RAYMOND,E
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: JOSEPH,NELLY
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 12/02/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A4
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 19:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:58

SUBJECT'S LAST NAME: REESE

DATE/TIME OF TEST: 01/03/2016 20:39:08

Diagnostics Check		OK	20:32
Air Blank		0.000	20:33
Control Test		0.080	20:33
Air Blank		0.000	20:34
Subject Sample	# 1	0.159	20:34
Breath Volume		2.652	20:34
Air Blank		0.000	20:35
Air Blank		0.000	20:37
Subject Sample	# 2	0.155	20:37
Breath Volume		2.261	20:37
Air Blank		0.000	20:38
Control Test		0.080	20:38
Air Blank		0.000	20:39
Diagnostics Check		OK	20:39

Data Download Date/Time: 02/02/2016 08:09

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002989

INSTRUMENT REGISTERED TO: TALLAHASSEE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHULZE,JON,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: SCHULZE,JON
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 12/31/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:56

SUBJECT'S LAST NAME: GRADDY

DATE/TIME OF TEST: 01/10/2016 03:23:20

Diagnostics Check		OK	03:16
Air Blank		0.000	03:16
Control Test		0.079	03:17
Air Blank		0.000	03:17
Subject Sample	# 1	0.122	03:18
Breath Volume		1.820	03:18
Air Blank		0.000	03:18
Air Blank		0.000	03:20
Subject Sample	# 2	0.133	03:21
Breath Volume		1.839	03:21
Air Blank		0.000	03:22
Control Test		0.079	03:22
Air Blank		0.000	03:23
Diagnostics Check		OK	03:23

Data Download Date/Time: 02/01/2016 08:31

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003224

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUASTO,NICHOLAS,
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: FERNANDEZ,JAVIER
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 12/14/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482986
DRY GAS EXPIRY DATE: 04/03/2016
VIOLATION CODE: DUI
ARREST TIME: 00:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: SHEKHOVTSOVA

DATE/TIME OF TEST: 01/23/2016 01:25:12

Diagnostics Check		OK	01:16
Air Blank		0.000	01:16
Control Test		0.080	01:16
Air Blank		0.000	01:17
Subject Sample	# 1	0.084	01:19
Breath Volume		1.363	01:19
Air Blank		0.000	01:19
Air Blank		0.000	01:21
Subject Sample	# 2	0.091	01:23
Breath Volume		1.269	01:23
Air Blank		0.000	01:24
Control Test		0.079	01:24
Air Blank		0.000	01:25
Diagnostics Check		OK	01:25

Data Download Date/Time: 01/26/2016 10:42

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003224

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SOCARRAS,PEDRO,A
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: MILLAN,KEVIN
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 12/14/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482986
DRY GAS EXPIRY DATE: 04/03/2016
VIOLATION CODE: DUI
ARREST TIME: 04:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:28

SUBJECT'S LAST NAME: LODEDO

DATE/TIME OF TEST: 01/23/2016 04:55:58

Diagnostics Check		OK	04:49
Air Blank		0.000	04:49
Control Test		0.080	04:49
Air Blank		0.000	04:50
Subject Sample	# 1	0.107	04:51
Breath Volume		1.921	04:51
Air Blank		0.000	04:51
Air Blank		0.000	04:53
Subject Sample	# 2	0.110	04:54
Breath Volume		2.410	04:54
Air Blank		0.000	04:55
Control Test		0.079	04:55
Air Blank		0.000	04:55
Diagnostics Check		OK	04:55

Data Download Date/Time: 01/26/2016 10:42

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003224

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SOCARRAS,PEDRO,A
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: FERNANDEZ,JAVIER
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 01/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482986
DRY GAS EXPIRY DATE: 04/03/2016
VIOLATION CODE: DUI
ARREST TIME: 02:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:27

SUBJECT'S LAST NAME: GARCON

DATE/TIME OF TEST: 01/31/2016 02:53:53

Diagnostics Check		OK	02:50	
Air Blank		0.000	02:51	
Control Test		0.080	02:51	
Air Blank		0.000	02:52	
Subject Sample	# 1		02:52	Subject Test Refused
Breath Volume		0.000	02:52	
Air Blank		0.000	02:52	
Control Test		0.079	02:53	
Air Blank		0.000	02:53	
Diagnostics Check		OK	02:53	

Data Download Date/Time: 02/23/2016 10:42

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003347

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,ROBERT,R
BREATH TEST OPERATOR AGENCY: Port Richey PD
ARRESTING OFFICER: WILLIAMS,ROBERT
ARRESTING OFFICER AGENCY: Port Richey PD
LAST AGENCY INSPECTION: 01/19/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A1
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:40

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 01/21/2016 06:06:42

Diagnostics Check		OK	06:00
Air Blank		0.000	06:00
Control Test		0.078	06:01
Air Blank		0.000	06:01
Subject Sample	# 1	0.081	06:02
Breath Volume		1.554	06:02
Air Blank		0.000	06:02
Air Blank		0.000	06:04
Subject Sample	# 2	0.080	06:05
Breath Volume		1.660	06:05
Air Blank		0.000	06:05
Control Test		0.077	06:06
Air Blank		0.000	06:06
Diagnostics Check		OK	06:06

Data Download Date/Time: 02/25/2016 08:00

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003347

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORRIS,ADAM,M
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: MORRIS,ADAM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/19/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A1
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:45

SUBJECT'S LAST NAME: MCMICHAEL

DATE/TIME OF TEST: 01/22/2016 05:11:21

Diagnostics Check		OK	05:05
Air Blank		0.000	05:05
Control Test		0.078	05:06
Air Blank		0.000	05:06
Subject Sample	# 1	0.149	05:07
Breath Volume		2.187	05:07
Air Blank		0.000	05:07
Air Blank		0.000	05:09
Subject Sample	# 2	0.144	05:09
Breath Volume		2.160	05:09
Air Blank		0.000	05:10
Control Test		0.078	05:10
Air Blank		0.000	05:11
Diagnostics Check		OK	05:11

Data Download Date/Time: 02/25/2016 08:00

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003347

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: METZLER,SHANE,A
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: METZLER,SHANE
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 01/19/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A1
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:53

SUBJECT'S LAST NAME: MOORE

DATE/TIME OF TEST: 01/24/2016 02:28:08

Diagnostics Check		OK	02:13	
Air Blank		0.000	02:13	
Control Test		0.078	02:13	
Air Blank		0.000	02:14	
Subject Sample	# 1		02:17	No Sample Provided
Breath Volume		0.000	02:17	
Air Blank		0.000	02:17	
Air Blank		0.000	02:19	
Subject Sample	# 2	0.118	02:21	
Breath Volume		2.113	02:21	
Air Blank		0.000	02:22	
Air Blank		0.000	02:23	
Subject Sample	# 3	0.119	02:26	
Breath Volume		1.289	02:26	
Air Blank		0.000	02:27	
Control Test		0.077	02:27	
Air Blank		0.000	02:28	
Diagnostics Check		OK	02:28	

Data Download Date/Time: 02/25/2016 08:00

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003347

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KELLY,WESLEY,R
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: KELLY,WESLEY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/19/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A1
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:32

SUBJECT'S LAST NAME: RICE

DATE/TIME OF TEST: 01/24/2016 05:00:44

Diagnostics Check		OK	04:53
Air Blank		0.000	04:54
Control Test		0.078	04:54
Air Blank		0.000	04:55
Subject Sample	# 1	0.048	04:56
Breath Volume		1.562	04:56
Air Blank		0.000	04:56
Air Blank		0.000	04:58
Subject Sample	# 2	0.047	04:59
Breath Volume		1.593	04:59
Air Blank		0.000	04:59
Control Test		0.078	05:00
Air Blank		0.000	05:00
Diagnostics Check		OK	05:00

Data Download Date/Time: 02/25/2016 08:00

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003347

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREEN,WILLIAM,E
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: GREEN,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/19/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A1
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 17:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:16

SUBJECT'S LAST NAME: SUTTER

DATE/TIME OF TEST: 01/26/2016 19:44:22

Diagnostics Check		OK	19:38
Air Blank		0.000	19:38
Control Test		0.079	19:39
Air Blank		0.000	19:39
Subject Sample	# 1	0.000	19:40
Breath Volume		2.902	19:40
Air Blank		0.000	19:40
Air Blank		0.000	19:42
Subject Sample	# 2	0.000	19:43
Breath Volume		2.621	19:43
Air Blank		0.000	19:43
Control Test		0.078	19:43
Air Blank		0.000	19:44
Diagnostics Check		OK	19:44

Data Download Date/Time: 02/25/2016 08:00

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003347

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHUCK,SCOTT,M
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: SCHUCK,SCOTT
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 01/19/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A1
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:41

SUBJECT'S LAST NAME: BONELLI

DATE/TIME OF TEST: 01/29/2016 02:11:36

Diagnostics Check		OK	02:03
Air Blank		0.000	02:04
Control Test		0.078	02:04
Air Blank		0.000	02:05
Subject Sample	# 1	0.183	02:06
Breath Volume		2.535	02:06
Air Blank		0.000	02:07
Air Blank		0.000	02:09
Subject Sample	# 2	0.177	02:10
Breath Volume		1.937	02:10
Air Blank		0.000	02:10
Control Test		0.077	02:11
Air Blank		0.000	02:11
Diagnostics Check		OK	02:11

Data Download Date/Time: 02/25/2016 08:00

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003347

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NIXON,BARRY,M
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: DAUGHERTY,DANIEL
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 01/19/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A1
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:27

SUBJECT'S LAST NAME: ROBINSON

DATE/TIME OF TEST: 01/29/2016 23:50:52

Diagnostics Check		OK	23:48	
Air Blank		0.000	23:48	
Control Test		0.077	23:48	
Air Blank		0.000	23:49	
Subject Sample	# 1		23:49	Subject Test Refused
Breath Volume		0.000	23:49	
Air Blank		0.000	23:49	
Control Test		0.077	23:50	
Air Blank		0.000	23:50	
Diagnostics Check		OK	23:50	

Data Download Date/Time: 02/25/2016 08:00

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003347

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NIXON,BARRY,M
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: KANE,RICHARD
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 01/19/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A1
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: LAMP

DATE/TIME OF TEST: 01/30/2016 03:47:28

Diagnostics Check		OK	03:40
Air Blank		0.000	03:41
Control Test		0.078	03:41
Air Blank		0.000	03:42
Subject Sample	# 1	0.147	03:43
Breath Volume		1.121	03:43
Air Blank		0.000	03:43
Air Blank		0.000	03:45
Subject Sample	# 2	0.145	03:46
Breath Volume		2.433	03:46
Air Blank		0.000	03:46
Control Test		0.078	03:46
Air Blank		0.000	03:47
Diagnostics Check		OK	03:47

Data Download Date/Time: 02/25/2016 08:00

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003347

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOUTROS,PHILLIP,A
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: BOUTROS,PHILLIP
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/19/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A1
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 01/31/2016 00:36:19

Diagnostics Check		OK	00:24	
Air Blank		0.000	00:24	
Control Test		0.079	00:24	
Air Blank		0.000	00:25	
Subject Sample	# 1	0.261	00:25	
Breath Volume		2.468	00:25	
Air Blank		0.000	00:26	
Air Blank		0.000	00:28	
Subject Sample	# 2	0.227	00:31	Slope Not Level
Breath Volume		1.773	00:31	
Air Blank		0.000	00:32	
Air Blank		0.000	00:33	
Subject Sample	# 3	0.249	00:34	
Breath Volume		3.828	00:34	
Air Blank		0.000	00:35	
Control Test		0.076	00:35	
Air Blank		0.000	00:36	
Diagnostics Check		OK	00:36	

Data Download Date/Time: 02/25/2016 08:00

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003347

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,WILLIAM,T
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SMITH,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/19/2016

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 03415080A1
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:56

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 01/31/2016 05:24:57

Diagnostics Check		OK	05:18
Air Blank		0.000	05:19
Control Test		0.077	05:19
Air Blank		0.000	05:19
Subject Sample	# 1	0.123	05:20
Breath Volume		1.472	05:20
Air Blank		0.000	05:21
Air Blank		0.000	05:22
Subject Sample	# 2	0.130	05:23
Breath Volume		1.878	05:23
Air Blank		0.000	05:24
Control Test		0.078	05:24
Air Blank		0.000	05:24
Diagnostics Check		OK	05:24

Data Download Date/Time: 02/25/2016 08:00

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003348

INSTRUMENT REGISTERED TO: SWEETWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOADA,CHRISTIAN,
BREATH TEST OPERATOR AGENCY: Sweetwater PD
ARRESTING OFFICER: OCEGUERA,YASMANI
ARRESTING OFFICER AGENCY: Florida Intern Unv
LAST AGENCY INSPECTION: 01/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1694353
DRY GAS EXPIRY DATE: 06/10/2017
VIOLATION CODE: DUI
ARREST TIME: 04:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:31

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 01/31/2016 05:06:47

Diagnostics Check		OK	04:59
Air Blank		0.000	05:00
Control Test		0.080	05:00
Air Blank		0.000	05:01
Subject Sample	# 1	0.169	05:01
Breath Volume		1.261	05:01
Air Blank		0.000	05:02
Air Blank		0.000	05:03
Subject Sample	# 2	0.152	05:05
Breath Volume		1.183	05:05
Air Blank		0.000	05:05
Control Test		0.079	05:06
Air Blank		0.000	05:06
Diagnostics Check		OK	05:06

Data Download Date/Time: 02/27/2016 11:33

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GONZALEZ,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 12/10/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:28

SUBJECT'S LAST NAME: MANGOLD

DATE/TIME OF TEST: 01/01/2016 02:55:39

Diagnostics Check		OK	02:48
Air Blank		0.000	02:49
Control Test		0.078	02:49
Air Blank		0.000	02:49
Subject Sample	# 1	0.190	02:51
Breath Volume		2.671	02:51
Air Blank		0.000	02:51
Air Blank		0.000	02:53
Subject Sample	# 2	0.189	02:54
Breath Volume		1.527	02:54
Air Blank		0.000	02:54
Control Test		0.077	02:55
Air Blank		0.000	02:55
Diagnostics Check		OK	02:55

Data Download Date/Time: 01/04/2016 20:35

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY,PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 12/10/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 05:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:19

SUBJECT'S LAST NAME: RAVELO

DATE/TIME OF TEST: 01/01/2016 06:44:27

Diagnostics Check		OK	06:40	
Air Blank		0.000	06:40	
Control Test		0.079	06:41	
Air Blank		0.000	06:41	
Subject Sample	# 1		06:43	Subject Test Refused
Breath Volume		0.000	06:43	
Air Blank		0.000	06:43	
Control Test		0.079	06:43	
Air Blank		0.000	06:44	
Diagnostics Check		OK	06:44	

Data Download Date/Time: 01/04/2016 20:35

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARR,BARBARA,H
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WOODEN,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/10/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 06:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:53

SUBJECT'S LAST NAME: GUTIERREZ

DATE/TIME OF TEST: 01/01/2016 07:20:45

Diagnostics Check		OK	07:14
Air Blank		0.000	07:15
Control Test		0.078	07:15
Air Blank		0.000	07:15
Subject Sample	# 1	0.175	07:16
Breath Volume		2.113	07:16
Air Blank		0.000	07:17
Air Blank		0.000	07:18
Subject Sample	# 2	0.171	07:19
Breath Volume		2.015	07:19
Air Blank		0.000	07:19
Control Test		0.079	07:20
Air Blank		0.000	07:20
Diagnostics Check		OK	07:20

Data Download Date/Time: 01/04/2016 20:35

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 12/10/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:18

SUBJECT'S LAST NAME: PFEIFER

DATE/TIME OF TEST: 01/02/2016 00:50:08

Diagnostics Check		OK	00:38	
Air Blank		0.000	00:38	
Control Test		0.079	00:39	
Air Blank		0.000	00:39	
Subject Sample	# 1	0.178	00:42	Volume Not Met
Breath Volume		0.082	00:42	
Air Blank		0.000	00:43	
Air Blank		0.000	00:45	
Subject Sample	# 2	0.191	00:48	Volume Not Met
Breath Volume		0.687	00:48	
Air Blank		0.000	00:49	
Control Test		0.076	00:49	
Air Blank		0.000	00:50	
Diagnostics Check		OK	00:50	

Data Download Date/Time: 01/04/2016 20:35

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 12/10/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:18

SUBJECT'S LAST NAME: PFEIFER

DATE/TIME OF TEST: 01/02/2016 01:01:20

Diagnostics Check		OK	00:51	
Air Blank		0.000	00:52	
Control Test		0.076	00:52	
Air Blank		0.000	00:53	
Subject Sample	# 1	0.219	00:56	Volume Not Met
Breath Volume		0.300	00:56	
Air Blank		0.000	00:56	
Air Blank		0.000	00:58	
Subject Sample	# 2		00:59	Subject Test Refused
Breath Volume		0.000	00:59	
Air Blank		0.000	01:00	
Control Test		0.076	01:00	
Air Blank		0.000	01:01	
Diagnostics Check		OK	01:01	

Data Download Date/Time: 01/04/2016 20:35

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY,PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 12/10/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:29

SUBJECT'S LAST NAME: DERY JR

DATE/TIME OF TEST: 01/02/2016 03:56:02

Diagnostics Check		OK	03:49
Air Blank		0.000	03:50
Control Test		0.078	03:50
Air Blank		0.000	03:50
Subject Sample	# 1	0.122	03:51
Breath Volume		3.683	03:51
Air Blank		0.000	03:52
Air Blank		0.000	03:53
Subject Sample	# 2	0.124	03:54
Breath Volume		3.683	03:54
Air Blank		0.000	03:55
Control Test		0.077	03:55
Air Blank		0.000	03:55
Diagnostics Check		OK	03:56

Data Download Date/Time: 01/04/2016 20:35

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 12/10/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:28

SUBJECT'S LAST NAME: ROSS

DATE/TIME OF TEST: 01/03/2016 00:55:19

Diagnostics Check		OK	00:49
Air Blank		0.000	00:49
Control Test		0.079	00:49
Air Blank		0.000	00:50
Subject Sample	# 1	0.000	00:50
Breath Volume		4.261	00:50
Air Blank		0.000	00:51
Air Blank		0.000	00:53
Subject Sample	# 2	0.000	00:53
Breath Volume		1.582	00:53
Air Blank		0.000	00:54
Control Test		0.079	00:54
Air Blank		0.000	00:55
Diagnostics Check		OK	00:55

Data Download Date/Time: 01/04/2016 20:35

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WOODEN,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/10/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: BRYANT JR

DATE/TIME OF TEST: 01/03/2016 02:16:55

Diagnostics Check		OK	02:08
Air Blank		0.000	02:09
Control Test		0.078	02:09
Air Blank		0.000	02:10
Subject Sample	# 1	0.150	02:12
Breath Volume		1.136	02:12
Air Blank		0.000	02:13
Air Blank		0.000	02:14
Subject Sample	# 2	0.157	02:15
Breath Volume		1.226	02:15
Air Blank		0.000	02:16
Control Test		0.077	02:16
Air Blank		0.000	02:16
Diagnostics Check		OK	02:16

Data Download Date/Time: 01/04/2016 20:35

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK,CASEY,S
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GONZALEZ,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 12/10/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:54

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 01/03/2016 04:23:14

Diagnostics Check		OK	04:16
Air Blank		0.000	04:16
Control Test		0.079	04:16
Air Blank		0.000	04:17
Subject Sample	# 1	0.123	04:18
Breath Volume		2.652	04:18
Air Blank		0.000	04:19
Air Blank		0.000	04:20
Subject Sample	# 2	0.117	04:21
Breath Volume		1.640	04:21
Air Blank		0.000	04:22
Control Test		0.078	04:22
Air Blank		0.000	04:23
Diagnostics Check		OK	04:23

Data Download Date/Time: 01/04/2016 20:35

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK,CASEY,S
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 12/10/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:37

SUBJECT'S LAST NAME: DOUGLAS

DATE/TIME OF TEST: 01/04/2016 03:03:53

Diagnostics Check		OK	02:57
Air Blank		0.000	02:58
Control Test		0.079	02:58
Air Blank		0.000	02:58
Subject Sample	# 1	0.102	02:59
Breath Volume		2.726	02:59
Air Blank		0.000	03:00
Air Blank		0.000	03:01
Subject Sample	# 2	0.102	03:02
Breath Volume		2.683	03:02
Air Blank		0.000	03:02
Control Test		0.079	03:03
Air Blank		0.000	03:03
Diagnostics Check		OK	03:03

Data Download Date/Time: 01/04/2016 20:35

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK,CASEY,S
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:07

SUBJECT'S LAST NAME: WICKS

DATE/TIME OF TEST: 01/05/2016 04:34:25

Diagnostics Check		OK	04:27
Air Blank		0.000	04:28
Control Test		0.080	04:28
Air Blank		0.000	04:29
Subject Sample	# 1	0.179	04:29
Breath Volume		3.476	04:29
Air Blank		0.000	04:30
Air Blank		0.000	04:32
Subject Sample	# 2	0.186	04:32
Breath Volume		2.679	04:32
Air Blank		0.000	04:33
Control Test		0.079	04:33
Air Blank		0.000	04:34
Diagnostics Check		OK	04:34

Data Download Date/Time: 02/01/2016 21:06

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARLAUG,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:49

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 01/07/2016 02:15:54

Diagnostics Check		OK	02:09
Air Blank		0.000	02:09
Control Test		0.078	02:10
Air Blank		0.000	02:10
Subject Sample	# 1	0.188	02:11
Breath Volume		3.156	02:11
Air Blank		0.000	02:11
Air Blank		0.000	02:13
Subject Sample	# 2	0.194	02:14
Breath Volume		4.019	02:14
Air Blank		0.000	02:14
Control Test		0.078	02:15
Air Blank		0.000	02:15
Diagnostics Check		OK	02:15

Data Download Date/Time: 02/01/2016 21:06

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK,CASEY,S
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:59

SUBJECT'S LAST NAME: PILE

DATE/TIME OF TEST: 01/08/2016 00:25:43

Diagnostics Check		OK	00:19
Air Blank		0.000	00:20
Control Test		0.079	00:20
Air Blank		0.000	00:20
Subject Sample	# 1	0.198	00:21
Breath Volume		1.699	00:21
Air Blank		0.000	00:21
Air Blank		0.000	00:23
Subject Sample	# 2	0.200	00:24
Breath Volume		1.539	00:24
Air Blank		0.000	00:24
Control Test		0.079	00:25
Air Blank		0.000	00:25
Diagnostics Check		OK	00:25

Data Download Date/Time: 02/01/2016 21:06

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK,CASEY,S
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GONZALEZ,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: LAZO ABREU

DATE/TIME OF TEST: 01/09/2016 02:27:11

Diagnostics Check		OK	02:20
Air Blank		0.000	02:20
Control Test		0.079	02:20
Air Blank		0.000	02:21
Subject Sample	# 1	0.108	02:22
Breath Volume		1.589	02:22
Air Blank		0.000	02:23
Air Blank		0.000	02:25
Subject Sample	# 2	0.105	02:25
Breath Volume		1.199	02:25
Air Blank		0.000	02:26
Control Test		0.078	02:26
Air Blank		0.000	02:27
Diagnostics Check		OK	02:27

Data Download Date/Time: 02/01/2016 21:06

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SARRO,TAYLOR,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GOFF,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:46

SUBJECT'S LAST NAME: RENDON-CALZADA

DATE/TIME OF TEST: 01/10/2016 04:13:14

Diagnostics Check		OK	04:06
Air Blank		0.000	04:07
Control Test		0.079	04:07
Air Blank		0.000	04:08
Subject Sample	# 1	0.111	04:08
Breath Volume		1.375	04:08
Air Blank		0.000	04:09
Air Blank		0.000	04:11
Subject Sample	# 2	0.108	04:11
Breath Volume		1.417	04:11
Air Blank		0.000	04:12
Control Test		0.078	04:12
Air Blank		0.000	04:13
Diagnostics Check		OK	04:13

Data Download Date/Time: 02/01/2016 21:06

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY,PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HOBSON,D
ARRESTING OFFICER AGENCY: U of So FL Campus PD
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:46

SUBJECT'S LAST NAME: NGUYEN

DATE/TIME OF TEST: 01/10/2016 05:15:11

Diagnostics Check		OK	05:08
Air Blank		0.000	05:09
Control Test		0.079	05:09
Air Blank		0.000	05:10
Subject Sample	# 1	0.177	05:10
Breath Volume		1.812	05:10
Air Blank		0.000	05:11
Air Blank		0.000	05:13
Subject Sample	# 2	0.171	05:13
Breath Volume		1.542	05:13
Air Blank		0.000	05:14
Control Test		0.078	05:14
Air Blank		0.000	05:15
Diagnostics Check		OK	05:15

Data Download Date/Time: 02/01/2016 21:06

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY,PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DAWSY,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:04

SUBJECT'S LAST NAME: HOSKINSON

DATE/TIME OF TEST: 01/11/2016 02:31:08

Diagnostics Check		OK	02:24
Air Blank		0.000	02:25
Control Test		0.079	02:25
Air Blank		0.000	02:26
Subject Sample	# 1	0.157	02:26
Breath Volume		1.472	02:26
Air Blank		0.000	02:27
Air Blank		0.000	02:29
Subject Sample	# 2	0.162	02:29
Breath Volume		2.183	02:29
Air Blank		0.000	02:30
Control Test		0.079	02:30
Air Blank		0.000	02:31
Diagnostics Check		OK	02:31

Data Download Date/Time: 02/01/2016 21:06

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HERNANDEZ,A
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:38

SUBJECT'S LAST NAME: TIMMONS

DATE/TIME OF TEST: 01/12/2016 03:07:09

Diagnostics Check		OK	02:58
Air Blank		0.000	02:59
Control Test		0.079	02:59
Air Blank		0.000	03:00
Subject Sample	# 1	0.019	03:01
Breath Volume		1.339	03:01
Air Blank		0.000	03:01
Air Blank		0.000	03:03
Subject Sample	# 2	0.020	03:05
Breath Volume		1.109	03:05
Air Blank		0.000	03:06
Control Test		0.081	03:06
Air Blank		0.000	03:07
Diagnostics Check		OK	03:07

Data Download Date/Time: 02/01/2016 21:06

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK,CASEY,S
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DAWSY,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:13

SUBJECT'S LAST NAME: ROGERS

DATE/TIME OF TEST: 01/14/2016 02:41:40

Diagnostics Check		OK	02:33
Air Blank		0.000	02:34
Control Test		0.080	02:34
Air Blank		0.000	02:34
Subject Sample	# 1	0.176	02:37
Breath Volume		2.414	02:37
Air Blank		0.000	02:37
Air Blank		0.000	02:39
Subject Sample	# 2	0.178	02:40
Breath Volume		1.429	02:40
Air Blank		0.000	02:40
Control Test		0.078	02:41
Air Blank		0.000	02:41
Diagnostics Check		OK	02:41

Data Download Date/Time: 02/01/2016 21:06

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARR,BARBARA,H
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAKER,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 16:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:04

SUBJECT'S LAST NAME: MUNOZ

DATE/TIME OF TEST: 01/15/2016 18:30:43

Diagnostics Check		OK	18:24
Air Blank		0.000	18:25
Control Test		0.078	18:25
Air Blank		0.000	18:25
Subject Sample	# 1	0.053	18:26
Breath Volume		2.039	18:26
Air Blank		0.000	18:27
Air Blank		0.000	18:28
Subject Sample	# 2	0.048	18:29
Breath Volume		1.523	18:29
Air Blank		0.000	18:29
Control Test		0.078	18:30
Air Blank		0.000	18:30
Diagnostics Check		OK	18:30

Data Download Date/Time: 02/01/2016 21:06

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SARRO,TAYLOR,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:32

SUBJECT'S LAST NAME: REINDEL

DATE/TIME OF TEST: 01/16/2016 01:01:03

Diagnostics Check		OK	00:54
Air Blank		0.000	00:54
Control Test		0.079	00:55
Air Blank		0.000	00:55
Subject Sample	# 1	0.307	00:56
Breath Volume		2.230	00:56
Air Blank		0.000	00:57
Air Blank		0.000	00:59
Subject Sample	# 2	0.299	00:59
Breath Volume		2.589	00:59
Air Blank		0.000	01:00
Control Test		0.077	01:00
Air Blank		0.000	01:00
Diagnostics Check		OK	01:01

Data Download Date/Time: 02/01/2016 21:06

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BLANCHARD,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:39

SUBJECT'S LAST NAME: BAILEY

DATE/TIME OF TEST: 01/16/2016 02:06:51

Diagnostics Check		OK	02:00
Air Blank		0.000	02:00
Control Test		0.079	02:01
Air Blank		0.000	02:01
Subject Sample	# 1	0.079	02:02
Breath Volume		1.921	02:02
Air Blank		0.000	02:03
Air Blank		0.000	02:04
Subject Sample	# 2	0.083	02:05
Breath Volume		1.605	02:05
Air Blank		0.000	02:05
Control Test		0.078	02:06
Air Blank		0.000	02:06
Diagnostics Check		OK	02:06

Data Download Date/Time: 02/01/2016 21:06

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK,CASEY,S
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HERNANDEZ,A
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:39

SUBJECT'S LAST NAME: ISIDRO

DATE/TIME OF TEST: 01/17/2016 05:06:32

Diagnostics Check		OK	04:59
Air Blank		0.000	05:00
Control Test		0.079	05:00
Air Blank		0.000	05:01
Subject Sample	# 1	0.196	05:01
Breath Volume		1.316	05:01
Air Blank		0.000	05:02
Air Blank		0.000	05:04
Subject Sample	# 2	0.187	05:04
Breath Volume		1.808	05:04
Air Blank		0.000	05:05
Control Test		0.078	05:05
Air Blank		0.000	05:06
Diagnostics Check		OK	05:06

Data Download Date/Time: 02/01/2016 21:06

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK,CASEY,S
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MALONE,C
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:29

SUBJECT'S LAST NAME: SOLIS

DATE/TIME OF TEST: 01/17/2016 22:56:24

Diagnostics Check		OK	22:49
Air Blank		0.000	22:50
Control Test		0.079	22:50
Air Blank		0.000	22:50
Subject Sample	# 1	0.000	22:52
Breath Volume		2.554	22:52
Air Blank		0.000	22:52
Air Blank		0.000	22:54
Subject Sample	# 2	0.000	22:55
Breath Volume		2.273	22:55
Air Blank		0.000	22:55
Control Test		0.078	22:55
Air Blank		0.000	22:56
Diagnostics Check		OK	22:56

Data Download Date/Time: 02/01/2016 21:06

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MALONE,C
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: LESTER

DATE/TIME OF TEST: 01/18/2016 02:57:12

Diagnostics Check		OK	02:50
Air Blank		0.000	02:50
Control Test		0.079	02:51
Air Blank		0.000	02:51
Subject Sample	# 1	0.173	02:52
Breath Volume		3.949	02:52
Air Blank		0.000	02:53
Air Blank		0.000	02:54
Subject Sample	# 2	0.174	02:55
Breath Volume		3.671	02:55
Air Blank		0.000	02:56
Control Test		0.079	02:56
Air Blank		0.000	02:57
Diagnostics Check		OK	02:57

Data Download Date/Time: 02/01/2016 21:06

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GONZALEZ,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: CONRADSON

DATE/TIME OF TEST: 01/19/2016 03:41:43

Diagnostics Check		OK	03:35
Air Blank		0.000	03:35
Control Test		0.080	03:36
Air Blank		0.000	03:36
Subject Sample	# 1	0.195	03:37
Breath Volume		2.558	03:37
Air Blank		0.000	03:37
Air Blank		0.000	03:39
Subject Sample	# 2	0.197	03:40
Breath Volume		2.523	03:40
Air Blank		0.000	03:40
Control Test		0.079	03:41
Air Blank		0.000	03:41
Diagnostics Check		OK	03:41

Data Download Date/Time: 02/01/2016 21:06

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK,CASEY,S
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARLAUG,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:59

SUBJECT'S LAST NAME: MASSON

DATE/TIME OF TEST: 01/22/2016 02:25:37

Diagnostics Check		OK	02:19
Air Blank		0.000	02:19
Control Test		0.080	02:20
Air Blank		0.000	02:20
Subject Sample	# 1	0.000	02:21
Breath Volume		1.488	02:21
Air Blank		0.000	02:21
Air Blank		0.000	02:23
Subject Sample	# 2	0.000	02:24
Breath Volume		1.484	02:24
Air Blank		0.000	02:24
Control Test		0.080	02:25
Air Blank		0.000	02:25
Diagnostics Check		OK	02:25

Data Download Date/Time: 02/01/2016 21:06

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK,CASEY,S
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARLAUG,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:32

SUBJECT'S LAST NAME: HIGGINBOTHAM

DATE/TIME OF TEST: 01/23/2016 00:58:37

Diagnostics Check		OK	00:52
Air Blank		0.000	00:52
Control Test		0.079	00:53
Air Blank		0.000	00:53
Subject Sample	# 1	0.073	00:54
Breath Volume		3.855	00:54
Air Blank		0.000	00:54
Air Blank		0.000	00:56
Subject Sample	# 2	0.076	00:57
Breath Volume		3.460	00:57
Air Blank		0.000	00:57
Control Test		0.079	00:58
Air Blank		0.000	00:58
Diagnostics Check		OK	00:58

Data Download Date/Time: 02/01/2016 21:06

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK,CASEY,S
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:06

SUBJECT'S LAST NAME: GRAHAM

DATE/TIME OF TEST: 01/23/2016 03:36:52

Diagnostics Check		OK	03:26
Air Blank		0.000	03:26
Control Test		0.079	03:27
Air Blank		0.000	03:27
Subject Sample	# 1	0.133	03:29
Breath Volume		2.312	03:29
Air Blank		0.000	03:29
Air Blank		0.000	03:31
Subject Sample	# 2	0.109	03:32
Breath Volume		1.179	03:32
Air Blank		0.000	03:33
Air Blank		0.000	03:34
Subject Sample	# 3	0.124	03:35
Breath Volume		3.281	03:35
Air Blank		0.000	03:35
Control Test		0.078	03:36
Air Blank		0.000	03:36
Diagnostics Check		OK	03:36

Data Download Date/Time: 02/01/2016 21:06

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK,CASEY,S
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MINAYA,J
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:32

SUBJECT'S LAST NAME: GRAFFRATH

DATE/TIME OF TEST: 01/23/2016 04:04:21

Diagnostics Check		OK	03:53
Air Blank		0.000	03:53
Control Test		0.079	03:53
Air Blank		0.000	03:54
Subject Sample	# 1	0.145	03:57
Breath Volume		2.054	03:57
Air Blank		0.000	03:58
Air Blank		0.000	03:59
Subject Sample	# 2	0.135	04:02
Breath Volume		1.449	04:02
Air Blank		0.000	04:03
Control Test		0.077	04:03
Air Blank		0.000	04:04
Diagnostics Check		OK	04:04

Data Download Date/Time: 02/01/2016 21:06

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 19:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:25

SUBJECT'S LAST NAME: SHAFFER

DATE/TIME OF TEST: 01/23/2016 19:52:29

Diagnostics Check		OK	19:46
Air Blank		0.000	19:46
Control Test		0.079	19:46
Air Blank		0.000	19:47
Subject Sample	# 1	0.153	19:47
Breath Volume		2.058	19:47
Air Blank		0.000	19:48
Air Blank		0.000	19:50
Subject Sample	# 2	0.162	19:50
Breath Volume		2.007	19:50
Air Blank		0.000	19:51
Control Test		0.079	19:51
Air Blank		0.000	19:52
Diagnostics Check		OK	19:52

Data Download Date/Time: 02/01/2016 21:06

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SARRO,TAYLOR,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FRANKLIN,R
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:17

SUBJECT'S LAST NAME: CONNOR

DATE/TIME OF TEST: 01/24/2016 01:44:02

Diagnostics Check		OK	01:37
Air Blank		0.000	01:37
Control Test		0.079	01:38
Air Blank		0.000	01:38
Subject Sample	# 1	0.161	01:39
Breath Volume		1.375	01:39
Air Blank		0.000	01:40
Air Blank		0.000	01:41
Subject Sample	# 2	0.158	01:42
Breath Volume		1.640	01:42
Air Blank		0.000	01:43
Control Test		0.079	01:43
Air Blank		0.000	01:43
Diagnostics Check		OK	01:44

Data Download Date/Time: 02/01/2016 21:06

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY,PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRENNAN,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:33

SUBJECT'S LAST NAME: DIONNE

DATE/TIME OF TEST: 01/24/2016 03:00:19

Diagnostics Check		OK	02:53
Air Blank		0.000	02:54
Control Test		0.079	02:54
Air Blank		0.000	02:55
Subject Sample	# 1	0.200	02:55
Breath Volume		1.871	02:55
Air Blank		0.000	02:56
Air Blank		0.000	02:58
Subject Sample	# 2	0.200	02:58
Breath Volume		1.492	02:58
Air Blank		0.000	02:59
Control Test		0.079	02:59
Air Blank		0.000	03:00
Diagnostics Check		OK	03:00

Data Download Date/Time: 02/01/2016 21:06

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SARRO,TAYLOR,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:59

SUBJECT'S LAST NAME: OMBA

DATE/TIME OF TEST: 01/24/2016 03:27:43

Diagnostics Check		OK	03:19
Air Blank		0.000	03:20
Control Test		0.078	03:20
Air Blank		0.000	03:21
Subject Sample	# 1	0.170	03:23
Breath Volume		2.359	03:23
Air Blank		0.000	03:23
Air Blank		0.000	03:25
Subject Sample	# 2	0.166	03:26
Breath Volume		2.316	03:26
Air Blank		0.000	03:26
Control Test		0.077	03:27
Air Blank		0.000	03:27
Diagnostics Check		OK	03:27

Data Download Date/Time: 02/01/2016 21:06

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GONZALEZ,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:53

SUBJECT'S LAST NAME: MCKEEFREY

DATE/TIME OF TEST: 01/24/2016 05:20:19

Diagnostics Check		OK	05:13
Air Blank		0.000	05:14
Control Test		0.079	05:14
Air Blank		0.000	05:14
Subject Sample	# 1	0.084	05:15
Breath Volume		1.917	05:15
Air Blank		0.000	05:16
Air Blank		0.000	05:17
Subject Sample	# 2	0.082	05:18
Breath Volume		2.421	05:18
Air Blank		0.000	05:19
Control Test		0.080	05:19
Air Blank		0.000	05:20
Diagnostics Check		OK	05:20

Data Download Date/Time: 02/01/2016 21:06

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DAWSY,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:29

SUBJECT'S LAST NAME: PROBST

DATE/TIME OF TEST: 01/24/2016 23:56:13

Diagnostics Check		OK	23:49
Air Blank		0.000	23:50
Control Test		0.080	23:50
Air Blank		0.000	23:50
Subject Sample	# 1	0.114	23:51
Breath Volume		1.140	23:51
Air Blank		0.000	23:52
Air Blank		0.000	23:54
Subject Sample	# 2	0.134	23:54
Breath Volume		1.855	23:54
Air Blank		0.000	23:55
Control Test		0.079	23:55
Air Blank		0.000	23:56
Diagnostics Check		OK	23:56

Data Download Date/Time: 02/01/2016 21:06

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SARRO,TAYLOR,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: NIEUWENDAM,R
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: FOWLER

DATE/TIME OF TEST: 01/26/2016 04:23:19

Diagnostics Check		OK	04:20	
Air Blank		0.000	04:20	
Control Test		0.079	04:21	
Air Blank		0.000	04:21	
Subject Sample	# 1		04:21	Subject Test Refused
Breath Volume		0.000	04:21	
Air Blank		0.000	04:22	
Control Test		0.080	04:22	
Air Blank		0.000	04:23	
Diagnostics Check		OK	04:23	

Data Download Date/Time: 02/01/2016 21:06

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK,CASEY,S
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: AGUAYO,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:08

SUBJECT'S LAST NAME: RAMIREZ-VAZQUEZ

DATE/TIME OF TEST: 01/26/2016 21:38:01

Diagnostics Check		OK	21:30
Air Blank		0.000	21:31
Control Test		0.078	21:31
Air Blank		0.000	21:31
Subject Sample	# 1	0.105	21:33
Breath Volume		1.628	21:33
Air Blank		0.000	21:33
Air Blank		0.000	21:35
Subject Sample	# 2	0.109	21:36
Breath Volume		1.980	21:36
Air Blank		0.000	21:37
Control Test		0.078	21:37
Air Blank		0.000	21:37
Diagnostics Check		OK	21:38

Data Download Date/Time: 02/01/2016 21:06

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNIDER,P
ARRESTING OFFICER AGENCY: Plant City PD
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 19:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:19

SUBJECT'S LAST NAME: DARLING

DATE/TIME OF TEST: 01/27/2016 20:45:41

Diagnostics Check		OK	20:39
Air Blank		0.000	20:39
Control Test		0.079	20:40
Air Blank		0.000	20:40
Subject Sample	# 1	0.065	20:41
Breath Volume		3.218	20:41
Air Blank		0.000	20:41
Air Blank		0.000	20:43
Subject Sample	# 2	0.060	20:44
Breath Volume		2.011	20:44
Air Blank		0.000	20:44
Control Test		0.079	20:45
Air Blank		0.000	20:45
Diagnostics Check		OK	20:45

Data Download Date/Time: 02/01/2016 21:06

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LYON,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:34

SUBJECT'S LAST NAME: ELDER

DATE/TIME OF TEST: 01/29/2016 04:01:00

Diagnostics Check		OK	03:54
Air Blank		0.000	03:54
Control Test		0.079	03:55
Air Blank		0.000	03:55
Subject Sample	# 1	0.202	03:56
Breath Volume		3.632	03:56
Air Blank		0.000	03:57
Air Blank		0.000	03:58
Subject Sample	# 2	0.184	03:59
Breath Volume		1.593	03:59
Air Blank		0.000	04:00
Control Test		0.079	04:00
Air Blank		0.000	04:00
Diagnostics Check		OK	04:01

Data Download Date/Time: 02/01/2016 21:06

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY,PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:04

SUBJECT'S LAST NAME: PEREZ JR

DATE/TIME OF TEST: 01/30/2016 05:31:47

Diagnostics Check		OK	05:25
Air Blank		0.000	05:25
Control Test		0.079	05:25
Air Blank		0.000	05:26
Subject Sample	# 1	0.092	05:27
Breath Volume		1.476	05:27
Air Blank		0.000	05:27
Air Blank		0.000	05:29
Subject Sample	# 2	0.095	05:30
Breath Volume		1.925	05:30
Air Blank		0.000	05:30
Control Test		0.079	05:31
Air Blank		0.000	05:31
Diagnostics Check		OK	05:31

Data Download Date/Time: 02/01/2016 21:06

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK,CASEY,S
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MALONE,C
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: UNDERWOOD

DATE/TIME OF TEST: 01/30/2016 23:58:29

Diagnostics Check		OK	23:52
Air Blank		0.000	23:52
Control Test		0.079	23:53
Air Blank		0.000	23:53
Subject Sample	# 1	0.000	23:53
Breath Volume		1.144	23:53
Air Blank		0.000	23:54
Air Blank		0.000	23:56
Subject Sample	# 2	0.007	23:57
Breath Volume		1.671	23:57
Air Blank		0.000	23:57
Control Test		0.079	23:57
Air Blank		0.000	23:58
Diagnostics Check		OK	23:58

Data Download Date/Time: 02/01/2016 21:06

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK,CASEY,S
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MILLER,J
ARRESTING OFFICER AGENCY: Temple Terrace PD
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:22

SUBJECT'S LAST NAME: FELIX

DATE/TIME OF TEST: 01/31/2016 02:49:49

Diagnostics Check		OK	02:42
Air Blank		0.000	02:43
Control Test		0.079	02:43
Air Blank		0.000	02:44
Subject Sample	# 1	0.144	02:45
Breath Volume		2.554	02:45
Air Blank		0.000	02:46
Air Blank		0.000	02:47
Subject Sample	# 2	0.141	02:48
Breath Volume		2.734	02:48
Air Blank		0.000	02:48
Control Test		0.078	02:49
Air Blank		0.000	02:49
Diagnostics Check		OK	02:49

Data Download Date/Time: 02/01/2016 21:06

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SEQUEIRA,N
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:21

SUBJECT'S LAST NAME: QUINONES

DATE/TIME OF TEST: 01/31/2016 04:50:22

Diagnostics Check		OK	04:41
Air Blank		0.000	04:41
Control Test		0.079	04:42
Air Blank		0.000	04:42
Subject Sample	# 1	0.108	04:44
Breath Volume		1.300	04:44
Air Blank		0.000	04:45
Air Blank		0.000	04:47
Subject Sample	# 2	0.117	04:48
Breath Volume		2.390	04:48
Air Blank		0.000	04:49
Control Test		0.079	04:49
Air Blank		0.000	04:50
Diagnostics Check		OK	04:50

Data Download Date/Time: 02/01/2016 21:06

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003410

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZUBER,HEATHER,D
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: ZUBER,HEATHER
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 12/18/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07014080A2
DRY GAS EXPIRY DATE: 05/01/2016
VIOLATION CODE: DUI
ARREST TIME: 19:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:30

SUBJECT'S LAST NAME: FOSKEY

DATE/TIME OF TEST: 01/07/2016 20:58:14

Diagnostics Check		OK	20:51
Air Blank		0.000	20:52
Control Test		0.077	20:52
Air Blank		0.000	20:52
Subject Sample	# 1	0.192	20:53
Breath Volume		1.632	20:53
Air Blank		0.000	20:53
Air Blank		0.000	20:55
Subject Sample	# 2	0.185	20:56
Breath Volume		2.449	20:56
Air Blank		0.000	20:57
Control Test		0.080	20:57
Air Blank		0.000	20:58
Diagnostics Check		OK	20:58

Data Download Date/Time: 01/22/2016 18:39

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003935

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BEEKMAN,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: BEEKMAN,JOSHUA
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 01/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1578837
DRY GAS EXPIRY DATE: 10/16/2016
VIOLATION CODE: DUI
ARREST TIME: 19:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:51

SUBJECT'S LAST NAME: DEVLIN

DATE/TIME OF TEST: 01/08/2016 20:14:28

Diagnostics Check		OK	20:11	
Air Blank		0.000	20:11	
Control Test		0.080	20:11	
Air Blank		0.000	20:12	
Subject Sample	# 1		20:13	Subject Test Refused
Breath Volume		0.000	20:13	
Air Blank		0.000	20:13	
Control Test		0.080	20:13	
Air Blank		0.000	20:14	
Diagnostics Check		OK	20:14	

Data Download Date/Time: 02/08/2016 15:35

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003935

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BEEKMAN,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: BEEKMAN,JOSHUA
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 01/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1578837
DRY GAS EXPIRY DATE: 10/16/2016
VIOLATION CODE: DUI
ARREST TIME: 03:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:24

SUBJECT'S LAST NAME: RUIZ

DATE/TIME OF TEST: 01/09/2016 03:53:38

Diagnostics Check		OK	03:49	
Air Blank		0.000	03:50	
Control Test		0.080	03:50	
Air Blank		0.000	03:50	
Subject Sample	# 1		03:52	Subject Test Refused
Breath Volume		0.000	03:52	
Air Blank		0.000	03:52	
Control Test		0.079	03:53	
Air Blank		0.000	03:53	
Diagnostics Check		OK	03:53	

Data Download Date/Time: 02/08/2016 15:35

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003935

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BEEKMAN,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: BEEKMAN,JOSHUA
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 01/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1578837
DRY GAS EXPIRY DATE: 10/16/2016
VIOLATION CODE: DUI
ARREST TIME: 03:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:23

SUBJECT'S LAST NAME: FIGUEROA

DATE/TIME OF TEST: 01/10/2016 03:58:45

Diagnostics Check		OK	03:49
Air Blank		0.000	03:50
Control Test		0.080	03:50
Air Blank		0.000	03:51
Subject Sample	# 1	0.216	03:54
Breath Volume		3.179	03:54
Air Blank		0.000	03:55
Air Blank		0.000	03:56
Subject Sample	# 2	0.205	03:57
Breath Volume		2.265	03:57
Air Blank		0.000	03:57
Control Test		0.077	03:58
Air Blank		0.000	03:58
Diagnostics Check		OK	03:58

Data Download Date/Time: 02/08/2016 15:35

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003935

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BEEKMAN,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: BEEKMAN,JOSHUA
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 01/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1578837
DRY GAS EXPIRY DATE: 10/16/2016
VIOLATION CODE: DUI
ARREST TIME: 22:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:19

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 01/23/2016 22:45:31

Diagnostics Check		OK	22:39
Air Blank		0.000	22:39
Control Test		0.081	22:40
Air Blank		0.000	22:40
Subject Sample	# 1	0.146	22:41
Breath Volume		2.265	22:41
Air Blank		0.000	22:41
Air Blank		0.000	22:43
Subject Sample	# 2	0.155	22:44
Breath Volume		1.902	22:44
Air Blank		0.000	22:44
Control Test		0.079	22:44
Air Blank		0.000	22:45
Diagnostics Check		OK	22:45

Data Download Date/Time: 02/08/2016 15:35

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003935

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BEEKMAN,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: BEEKMAN,JOSHUA
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 01/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1578837
DRY GAS EXPIRY DATE: 10/16/2016
VIOLATION CODE: DUI
ARREST TIME: 03:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:39

SUBJECT'S LAST NAME: BUTLER

DATE/TIME OF TEST: 01/24/2016 04:12:33

Diagnostics Check		OK	04:03	
Air Blank		0.000	04:04	
Control Test		0.082	04:04	
Air Blank		0.000	04:05	
Subject Sample	# 1		04:08	No Sample Provided
Breath Volume		0.000	04:08	
Air Blank		0.000	04:08	
Air Blank		0.000	04:10	
Subject Sample	# 2		04:11	Subject Test Refused
Breath Volume		0.000	04:11	
Air Blank		0.000	04:11	
Control Test		0.081	04:11	
Air Blank		0.000	04:12	
Diagnostics Check		OK	04:12	

Data Download Date/Time: 02/08/2016 15:35

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003937

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.0.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEDULLA,VINCENT,A
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: PEDULLA,VINCENT
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 01/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1578837
DRY GAS EXPIRY DATE: 10/16/2016
VIOLATION CODE: DUI
ARREST TIME: 01:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:22

SUBJECT'S LAST NAME: CAUDILL

DATE/TIME OF TEST: 01/22/2016 01:26:41

Diagnostics Check		OK	01:16
Air Blank		0.000	01:17
Control Test		0.079	01:17
Air Blank		0.000	01:18
Subject Sample	# 1	0.214	01:18
Breath Volume		1.343	01:18
Air Blank		0.000	01:19
Air Blank		0.000	01:21
Subject Sample	# 2	0.237	01:21
Breath Volume		3.230	01:21
Air Blank		0.000	01:22
Air Blank		0.000	01:24
Subject Sample	# 3	0.209	01:25
Breath Volume		2.117	01:25
Air Blank		0.000	01:25
Control Test		0.078	01:26
Air Blank		0.000	01:26
Diagnostics Check		OK	01:26

Data Download Date/Time: 02/08/2016 15:38

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003937

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.0.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEDULLA,VINCENT,A
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: PEDULLA,VINCENT
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 01/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1578837
DRY GAS EXPIRY DATE: 10/16/2016
VIOLATION CODE: DUI
ARREST TIME: 19:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:30

SUBJECT'S LAST NAME: AYLOR

DATE/TIME OF TEST: 01/25/2016 20:13:22

Diagnostics Check		OK	20:06
Air Blank		0.000	20:07
Control Test		0.081	20:07
Air Blank		0.000	20:07
Subject Sample	# 1	0.245	20:08
Breath Volume		2.070	20:08
Air Blank		0.000	20:09
Air Blank		0.000	20:11
Subject Sample	# 2	0.240	20:11
Breath Volume		1.550	20:11
Air Blank		0.000	20:12
Control Test		0.079	20:12
Air Blank		0.000	20:13
Diagnostics Check		OK	20:13

Data Download Date/Time: 02/08/2016 15:38

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003937

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.0.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEDULLA,VINCENT,A
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: PEDULLA,VINCENT
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 01/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1578837
DRY GAS EXPIRY DATE: 10/16/2016
VIOLATION CODE: DUI
ARREST TIME: 03:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 01/26/2016 03:24:54

Diagnostics Check		OK	03:10	
Air Blank		0.000	03:10	
Control Test		0.081	03:11	
Air Blank		0.000	03:11	
Subject Sample	# 1	0.000	03:12	
Breath Volume		1.414	03:12	
Air Blank		0.000	03:13	
Air Blank		0.000	03:15	
Subject Sample	# 2	0.000	03:18	Volume Not Met
Breath Volume		0.375	03:18	
Air Blank		0.000	03:18	
Air Blank		0.000	03:20	
Subject Sample	# 3	0.000	03:23	
Breath Volume		1.964	03:23	
Air Blank		0.000	03:23	
Control Test		0.079	03:24	
Air Blank		0.000	03:24	
Diagnostics Check		OK	03:24	

Data Download Date/Time: 02/08/2016 15:38

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THACKER,ROBERT,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: THACKER,ROBERT
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 12/17/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:36

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 01/01/2016 02:08:29

Diagnostics Check		OK	02:01
Air Blank		0.000	02:01
Control Test		0.079	02:02
Air Blank		0.000	02:02
Subject Sample	# 1	0.255	02:03
Breath Volume		2.910	02:03
Air Blank		0.000	02:04
Air Blank		0.000	02:06
Subject Sample	# 2	0.239	02:06
Breath Volume		2.652	02:06
Air Blank		0.000	02:07
Control Test		0.077	02:07
Air Blank		0.000	02:08
Diagnostics Check		OK	02:08

Data Download Date/Time: 01/23/2016 20:01

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPENCE,SCOTT,
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: SPENCE,SCOTT
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 12/17/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:47

SUBJECT'S LAST NAME: SIMMONS

DATE/TIME OF TEST: 01/01/2016 03:15:13

Diagnostics Check		OK	03:08
Air Blank		0.000	03:08
Control Test		0.079	03:09
Air Blank		0.000	03:09
Subject Sample	# 1	0.115	03:10
Breath Volume		2.789	03:10
Air Blank		0.000	03:11
Air Blank		0.000	03:12
Subject Sample	# 2	0.109	03:13
Breath Volume		3.316	03:13
Air Blank		0.000	03:14
Control Test		0.078	03:14
Air Blank		0.000	03:15
Diagnostics Check		OK	03:15

Data Download Date/Time: 01/23/2016 20:01

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCOY,RICHARD,G
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: MCCOY,RICHARD
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 12/17/2015

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:22

SUBJECT'S LAST NAME: SKINNER

DATE/TIME OF TEST: 01/01/2016 03:55:26

Diagnostics Check		OK	03:43	
Air Blank		0.000	03:44	
Control Test		0.079	03:44	
Air Blank		0.000	03:45	
Subject Sample	# 1	0.099	03:48	Volume Not Met
Breath Volume		0.105	03:48	
Air Blank		0.000	03:48	
Air Blank		0.000	03:50	
Subject Sample	# 2	0.106	03:53	Volume Not Met
Breath Volume		0.144	03:53	
Air Blank		0.000	03:54	
Control Test		0.079	03:54	
Air Blank		0.000	03:55	
Diagnostics Check		OK	03:55	

Data Download Date/Time: 01/23/2016 20:01

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCOY,RICHARD,G
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: MCCOY,RICHARD
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 12/17/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:22

SUBJECT'S LAST NAME: SKINNER

DATE/TIME OF TEST: 01/01/2016 04:05:49

Diagnostics Check		OK	03:57
Air Blank		0.000	03:57
Control Test		0.079	03:58
Air Blank		0.000	03:58
Subject Sample	# 1	0.102	04:01
Breath Volume		1.980	04:01
Air Blank		0.000	04:01
Air Blank		0.000	04:03
Subject Sample	# 2	0.104	04:04
Breath Volume		2.625	04:04
Air Blank		0.000	04:04
Control Test		0.079	04:05
Air Blank		0.000	04:05
Diagnostics Check		OK	04:05

Data Download Date/Time: 01/23/2016 20:01

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREIRA,JOSE,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: PEREIRA,JOSE
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 12/17/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: CARMICHAEL

DATE/TIME OF TEST: 01/02/2016 03:34:51

Diagnostics Check		OK	03:26	
Air Blank		0.000	03:26	
Control Test		0.078	03:26	
Air Blank		0.000	03:27	
Subject Sample	# 1		03:30	No Sample Provided
Breath Volume		0.000	03:30	
Air Blank		0.000	03:31	
Air Blank		0.000	03:33	
Subject Sample	# 2		03:33	Subject Test Refused
Breath Volume		0.000	03:33	
Air Blank		0.000	03:33	
Control Test		0.080	03:34	
Air Blank		0.000	03:34	
Diagnostics Check		OK	03:34	

Data Download Date/Time: 01/23/2016 20:01

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILSON,WILLIAM,M
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: WILSON,WILLIAM
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 12/17/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 13:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:58

SUBJECT'S LAST NAME: KAISER

DATE/TIME OF TEST: 01/02/2016 14:33:42

Diagnostics Check		OK	14:25
Air Blank		0.000	14:26
Control Test		0.081	14:26
Air Blank		0.000	14:26
Subject Sample	# 1	0.233	14:28
Breath Volume		1.480	14:28
Air Blank		0.000	14:28
Air Blank		0.000	14:30
Subject Sample	# 2	0.228	14:32
Breath Volume		1.324	14:32
Air Blank		0.000	14:32
Control Test		0.077	14:33
Air Blank		0.000	14:33
Diagnostics Check		OK	14:33

Data Download Date/Time: 01/23/2016 20:01

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALEXANDER,C,L
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: NOBLES,DARIUS
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 12/17/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 06:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:31

SUBJECT'S LAST NAME: LANE

DATE/TIME OF TEST: 01/03/2016 07:57:42

Diagnostics Check		OK	07:51
Air Blank		0.000	07:52
Control Test		0.080	07:52
Air Blank		0.000	07:52
Subject Sample	# 1	0.081	07:53
Breath Volume		1.898	07:53
Air Blank		0.000	07:53
Air Blank		0.000	07:55
Subject Sample	# 2	0.081	07:56
Breath Volume		2.093	07:56
Air Blank		0.000	07:56
Control Test		0.079	07:57
Air Blank		0.000	07:57
Diagnostics Check		OK	07:57

Data Download Date/Time: 01/23/2016 20:01

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FORDHAM,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: SIMPSON,ANDREW
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 12/17/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 08:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:25

SUBJECT'S LAST NAME: BETHANY

DATE/TIME OF TEST: 01/03/2016 09:53:47

Diagnostics Check		OK	09:46
Air Blank		0.000	09:47
Control Test		0.080	09:47
Air Blank		0.000	09:47
Subject Sample	# 1	0.115	09:49
Breath Volume		1.253	09:49
Air Blank		0.000	09:49
Air Blank		0.000	09:51
Subject Sample	# 2	0.119	09:52
Breath Volume		1.785	09:52
Air Blank		0.000	09:52
Control Test		0.080	09:53
Air Blank		0.000	09:53
Diagnostics Check		OK	09:53

Data Download Date/Time: 01/23/2016 20:01

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PETERSON,ANTHONY,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: BROWN,RAPHEAL
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 12/17/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 18:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:28

SUBJECT'S LAST NAME: CHIARITO

DATE/TIME OF TEST: 01/03/2016 19:55:46

Diagnostics Check		OK	19:48
Air Blank		0.000	19:49
Control Test		0.080	19:49
Air Blank		0.000	19:50
Subject Sample	# 1	0.277	19:50
Breath Volume		1.769	19:50
Air Blank		0.000	19:51
Air Blank		0.000	19:53
Subject Sample	# 2	0.281	19:54
Breath Volume		2.011	19:54
Air Blank		0.000	19:54
Control Test		0.080	19:55
Air Blank		0.000	19:55
Diagnostics Check		OK	19:55

Data Download Date/Time: 01/23/2016 20:01

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREIRA,JOSE,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: PEREIRA,JOSE
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 12/17/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: HARGIS

DATE/TIME OF TEST: 01/05/2016 03:32:23

Diagnostics Check		OK	03:25
Air Blank		0.000	03:26
Control Test		0.080	03:26
Air Blank		0.000	03:26
Subject Sample	# 1	0.147	03:27
Breath Volume		1.832	03:27
Air Blank		0.000	03:28
Air Blank		0.000	03:30
Subject Sample	# 2	0.137	03:30
Breath Volume		1.886	03:30
Air Blank		0.000	03:31
Control Test		0.080	03:31
Air Blank		0.000	03:32
Diagnostics Check		OK	03:32

Data Download Date/Time: 01/23/2016 20:01

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREIRA,JOSE,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: PEREIRA,JOSE
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 12/17/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: PORTER

DATE/TIME OF TEST: 01/06/2016 00:54:58

Diagnostics Check		OK	00:46	
Air Blank		0.000	00:46	
Control Test		0.080	00:47	
Air Blank		0.000	00:47	
Subject Sample	# 1		00:50	No Sample Provided
Breath Volume		0.000	00:50	
Air Blank		0.000	00:51	
Air Blank		0.000	00:53	
Subject Sample	# 2		00:53	Subject Test Refused
Breath Volume		0.000	00:53	
Air Blank		0.000	00:54	
Control Test		0.080	00:54	
Air Blank		0.000	00:54	
Diagnostics Check		OK	00:54	

Data Download Date/Time: 01/23/2016 20:01

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALEXANDER,C,L
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: ALEXANDER,C
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 12/17/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 11:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:40

SUBJECT'S LAST NAME: FLEMING

DATE/TIME OF TEST: 01/08/2016 12:07:45

Diagnostics Check		OK	12:01
Air Blank		0.000	12:02
Control Test		0.079	12:02
Air Blank		0.000	12:03
Subject Sample	# 1	0.000	12:03
Breath Volume		1.519	12:03
Air Blank		0.000	12:03
Air Blank		0.000	12:05
Subject Sample	# 2	0.000	12:06
Breath Volume		1.582	12:06
Air Blank		0.000	12:06
Control Test		0.079	12:07
Air Blank		0.000	12:07
Diagnostics Check		OK	12:07

Data Download Date/Time: 01/23/2016 20:01

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPEAKMAN,ROBERT,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: SPEAKMAN,ROBERT
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 12/17/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:23

SUBJECT'S LAST NAME: LYERLY

DATE/TIME OF TEST: 01/08/2016 23:50:51

Diagnostics Check		OK	23:45	
Air Blank		0.000	23:46	
Control Test		0.080	23:46	
Air Blank		0.000	23:47	
Subject Sample	# 1		23:49	Subject Test Refused
Breath Volume		0.000	23:49	
Air Blank		0.000	23:49	
Control Test		0.079	23:50	
Air Blank		0.000	23:50	
Diagnostics Check		OK	23:50	

Data Download Date/Time: 01/23/2016 20:01

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,JONATHAN,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: HERNANDEZ,JONATHAN
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 12/17/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:33

SUBJECT'S LAST NAME: WRANN

DATE/TIME OF TEST: 01/09/2016 03:02:24

Diagnostics Check		OK	02:57	
Air Blank		0.000	02:58	
Control Test		0.079	02:58	
Air Blank		0.000	02:59	
Subject Sample	# 1		03:01	Subject Test Refused
Breath Volume		0.000	03:01	
Air Blank		0.000	03:01	
Control Test		0.079	03:01	
Air Blank		0.000	03:02	
Diagnostics Check		OK	03:02	

Data Download Date/Time: 01/23/2016 20:01

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSS,JAMIESON,S
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: WILDE,MATTHEW
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 12/17/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 05:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:22

SUBJECT'S LAST NAME: GUTHRIE

DATE/TIME OF TEST: 01/09/2016 05:47:10

Diagnostics Check		OK	05:44	
Air Blank		0.000	05:44	
Control Test		0.080	05:45	
Air Blank		0.000	05:45	
Subject Sample	# 1		05:45	Subject Test Refused
Breath Volume		0.000	05:45	
Air Blank		0.000	05:46	
Control Test		0.079	05:46	
Air Blank		0.000	05:47	
Diagnostics Check		OK	05:47	

Data Download Date/Time: 01/23/2016 20:01

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PREJEAN,JESSICA,D
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: PREJEAN,JESSICA
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 12/17/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:24

SUBJECT'S LAST NAME: KUHN

DATE/TIME OF TEST: 01/09/2016 21:50:52

Diagnostics Check		OK	21:44
Air Blank		0.000	21:45
Control Test		0.079	21:45
Air Blank		0.000	21:45
Subject Sample	# 1	0.181	21:46
Breath Volume		3.437	21:46
Air Blank		0.000	21:47
Air Blank		0.000	21:48
Subject Sample	# 2	0.178	21:49
Breath Volume		3.550	21:49
Air Blank		0.000	21:49
Control Test		0.078	21:50
Air Blank		0.000	21:50
Diagnostics Check		OK	21:50

Data Download Date/Time: 01/23/2016 20:01

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PREJEAN,JESSICA,D
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: PREJEAN,JESSICA
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 12/17/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:17

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 01/10/2016 03:41:46

Diagnostics Check		OK	03:38	
Air Blank		0.000	03:38	
Control Test		0.080	03:39	
Air Blank		0.000	03:39	
Subject Sample	# 1		03:40	Subject Test Refused
Breath Volume		0.000	03:40	
Air Blank		0.000	03:40	
Control Test		0.080	03:41	
Air Blank		0.000	03:41	
Diagnostics Check		OK	03:41	

Data Download Date/Time: 01/23/2016 20:01

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LYBRAND,CHRISTOPHER,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: LYBRAND,CHRISTOPHER
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 12/17/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:22

SUBJECT'S LAST NAME: DOSTER

DATE/TIME OF TEST: 01/13/2016 00:48:55

Diagnostics Check		OK	00:42
Air Blank		0.000	00:43
Control Test		0.079	00:43
Air Blank		0.000	00:44
Subject Sample	# 1	0.172	00:44
Breath Volume		1.632	00:44
Air Blank		0.000	00:45
Air Blank		0.000	00:47
Subject Sample	# 2	0.177	00:47
Breath Volume		2.125	00:47
Air Blank		0.000	00:48
Control Test		0.079	00:48
Air Blank		0.000	00:48
Diagnostics Check		OK	00:48

Data Download Date/Time: 01/23/2016 20:01

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREIRA,JOSE,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: PEREIRA,JOSE
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 12/17/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:18

SUBJECT'S LAST NAME: YOST

DATE/TIME OF TEST: 01/14/2016 04:48:21

Diagnostics Check		OK	04:41
Air Blank		0.000	04:42
Control Test		0.079	04:42
Air Blank		0.000	04:43
Subject Sample	# 1	0.184	04:43
Breath Volume		1.539	04:43
Air Blank		0.000	04:44
Air Blank		0.000	04:46
Subject Sample	# 2	0.187	04:46
Breath Volume		1.511	04:46
Air Blank		0.000	04:47
Control Test		0.079	04:47
Air Blank		0.000	04:48
Diagnostics Check		OK	04:48

Data Download Date/Time: 01/23/2016 20:02

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREIRA,JOSE,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: PEREIRA,JOSE
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 12/17/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:16

SUBJECT'S LAST NAME: GLERUM

DATE/TIME OF TEST: 01/14/2016 22:46:27

Diagnostics Check		OK	22:37	
Air Blank		0.000	22:38	
Control Test		0.079	22:38	
Air Blank		0.000	22:39	
Subject Sample	# 1		22:42	No Sample Provided
Breath Volume		0.000	22:42	
Air Blank		0.000	22:42	
Air Blank		0.000	22:44	
Subject Sample	# 2		22:45	Subject Test Refused
Breath Volume		0.000	22:45	
Air Blank		0.000	22:45	
Control Test		0.081	22:45	
Air Blank		0.000	22:46	
Diagnostics Check		OK	22:46	

Data Download Date/Time: 01/23/2016 20:02

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBINSON,CHRISTOPHER,A
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: ROBINSON,CHRISTOPHER
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 12/17/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 05:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:30

SUBJECT'S LAST NAME: SOSA

DATE/TIME OF TEST: 01/15/2016 06:06:01

Diagnostics Check		OK	05:59
Air Blank		0.000	05:59
Control Test		0.080	06:00
Air Blank		0.000	06:00
Subject Sample	# 1	0.176	06:01
Breath Volume		2.050	06:01
Air Blank		0.000	06:02
Air Blank		0.000	06:03
Subject Sample	# 2	0.189	06:04
Breath Volume		1.199	06:04
Air Blank		0.000	06:05
Control Test		0.080	06:05
Air Blank		0.000	06:05
Diagnostics Check		OK	06:06

Data Download Date/Time: 01/23/2016 20:02

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLARK,KEITH,A
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: ROCKWELL,PETER
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 12/17/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: BUI
ARREST TIME: 16:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:03

SUBJECT'S LAST NAME: COULOMBE

DATE/TIME OF TEST: 01/15/2016 17:26:08

Diagnostics Check		OK	17:23	
Air Blank		0.000	17:23	
Control Test		0.079	17:24	
Air Blank		0.000	17:24	
Subject Sample	# 1		17:24	Subject Test Refused
Breath Volume		0.000	17:24	
Air Blank		0.000	17:25	
Control Test		0.080	17:25	
Air Blank		0.000	17:26	
Diagnostics Check		OK	17:26	

Data Download Date/Time: 01/23/2016 20:02

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,JONATHAN,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: HERNANDEZ,JONATHAN
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 12/17/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:52

SUBJECT'S LAST NAME: EUBANKS

DATE/TIME OF TEST: 01/16/2016 04:17:06

Diagnostics Check		OK	04:13	
Air Blank		0.000	04:14	
Control Test		0.080	04:14	
Air Blank		0.000	04:15	
Subject Sample	# 1		04:15	Subject Test Refused
Breath Volume		0.000	04:15	
Air Blank		0.000	04:16	
Control Test		0.079	04:16	
Air Blank		0.000	04:17	
Diagnostics Check		OK	04:17	

Data Download Date/Time: 01/23/2016 20:02

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PASCHAL,THOMAS,C
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: PASCHAL,THOMAS
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 12/17/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 05:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:10

SUBJECT'S LAST NAME: BROOKS III

DATE/TIME OF TEST: 01/17/2016 06:52:25

Diagnostics Check		OK	06:46
Air Blank		0.000	06:46
Control Test		0.080	06:46
Air Blank		0.000	06:47
Subject Sample	# 1	0.192	06:47
Breath Volume		2.796	06:47
Air Blank		0.000	06:48
Air Blank		0.000	06:50
Subject Sample	# 2	0.194	06:50
Breath Volume		3.667	06:50
Air Blank		0.000	06:51
Control Test		0.079	06:51
Air Blank		0.000	06:52
Diagnostics Check		OK	06:52

Data Download Date/Time: 01/23/2016 20:02

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LYBRAND,CHRISTOPHER,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: LYBRAND,CHRISTOPHER
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 12/17/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: POIALII

DATE/TIME OF TEST: 01/21/2016 01:36:56

Diagnostics Check		OK	01:30
Air Blank		0.000	01:31
Control Test		0.079	01:31
Air Blank		0.000	01:32
Subject Sample	# 1	0.113	01:32
Breath Volume		2.195	01:32
Air Blank		0.000	01:33
Air Blank		0.000	01:35
Subject Sample	# 2	0.103	01:35
Breath Volume		2.101	01:35
Air Blank		0.000	01:36
Control Test		0.079	01:36
Air Blank		0.000	01:36
Diagnostics Check		OK	01:36

Data Download Date/Time: 01/23/2016 20:02

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAN EPS,MICHAEL,S
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: VAN EPS,MICHAEL
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 01/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 19:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:45

SUBJECT'S LAST NAME: MARSHALL

DATE/TIME OF TEST: 01/23/2016 21:08:59

Diagnostics Check		OK	21:06	
Air Blank		0.000	21:06	
Control Test		0.080	21:06	
Air Blank		0.000	21:07	
Subject Sample	# 1		21:07	Subject Test Refused
Breath Volume		0.000	21:07	
Air Blank		0.000	21:08	
Control Test		0.081	21:08	
Air Blank		0.000	21:08	
Diagnostics Check		OK	21:08	

Data Download Date/Time: 02/16/2016 17:08

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,GLENN,H
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: HAND,GLENN
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 01/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:09

SUBJECT'S LAST NAME: CURRIE

DATE/TIME OF TEST: 01/24/2016 00:38:43

Diagnostics Check		OK	00:35	
Air Blank		0.000	00:36	
Control Test		0.080	00:36	
Air Blank		0.000	00:37	
Subject Sample	# 1		00:37	Subject Test Refused
Breath Volume		0.000	00:37	
Air Blank		0.000	00:37	
Control Test		0.080	00:38	
Air Blank		0.000	00:38	
Diagnostics Check		OK	00:38	

Data Download Date/Time: 02/16/2016 17:08

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,GLENN,H
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: HAND,GLENN
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 01/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:02

SUBJECT'S LAST NAME: FREELAND

DATE/TIME OF TEST: 01/24/2016 03:25:59

Diagnostics Check		OK	03:22	
Air Blank		0.000	03:23	
Control Test		0.080	03:23	
Air Blank		0.000	03:24	
Subject Sample	# 1		03:24	Subject Test Refused
Breath Volume		0.000	03:24	
Air Blank		0.000	03:25	
Control Test		0.081	03:25	
Air Blank		0.000	03:25	
Diagnostics Check		OK	03:25	

Data Download Date/Time: 02/16/2016 17:08

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREIRA,JOSE,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: PEREIRA,JOSE
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 01/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:51

SUBJECT'S LAST NAME: GAGLOYEVA

DATE/TIME OF TEST: 01/24/2016 04:22:07

Diagnostics Check		OK	04:13	
Air Blank		0.000	04:14	
Control Test		0.080	04:14	
Air Blank		0.000	04:14	
Subject Sample	# 1		04:18	No Sample Provided
Breath Volume		0.000	04:18	
Air Blank		0.000	04:18	
Air Blank		0.000	04:20	
Subject Sample	# 2		04:20	Subject Test Refused
Breath Volume		0.000	04:20	
Air Blank		0.000	04:21	
Control Test		0.080	04:21	
Air Blank		0.000	04:22	
Diagnostics Check		OK	04:22	

Data Download Date/Time: 02/16/2016 17:08

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LYBRAND,CHRISTOPHER,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: LYBRAND,CHRISTOPHER
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 01/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:07

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 01/26/2016 02:34:29

Diagnostics Check		OK	02:28
Air Blank		0.000	02:28
Control Test		0.078	02:28
Air Blank		0.000	02:29
Subject Sample	# 1	0.247	02:29
Breath Volume		1.398	02:29
Air Blank		0.000	02:30
Air Blank		0.000	02:32
Subject Sample	# 2	0.241	02:32
Breath Volume		2.359	02:32
Air Blank		0.000	02:33
Control Test		0.079	02:33
Air Blank		0.000	02:34
Diagnostics Check		OK	02:34

Data Download Date/Time: 02/16/2016 17:08

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORBITT,KYLE,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: FREDERICK,TAVISH
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 01/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 11:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:16

SUBJECT'S LAST NAME: HARDISON

DATE/TIME OF TEST: 01/27/2016 12:49:41

Diagnostics Check		OK	12:41
Air Blank		0.000	12:41
Control Test		0.079	12:41
Air Blank		0.000	12:42
Subject Sample	# 1	0.000	12:44
Breath Volume		2.679	12:44
Air Blank		0.000	12:45
Air Blank		0.000	12:47
Subject Sample	# 2	0.000	12:48
Breath Volume		1.164	12:48
Air Blank		0.000	12:48
Control Test		0.077	12:49
Air Blank		0.000	12:49
Diagnostics Check		OK	12:49

Data Download Date/Time: 02/16/2016 17:08

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREIRA,JOSE,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: PEREIRA,JOSE
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 01/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:12

SUBJECT'S LAST NAME: BOWER

DATE/TIME OF TEST: 01/27/2016 21:40:39

Diagnostics Check		OK	21:34
Air Blank		0.000	21:35
Control Test		0.079	21:35
Air Blank		0.000	21:35
Subject Sample	# 1	0.000	21:36
Breath Volume		1.589	21:36
Air Blank		0.000	21:36
Air Blank		0.000	21:38
Subject Sample	# 2	0.000	21:39
Breath Volume		1.218	21:39
Air Blank		0.000	21:39
Control Test		0.080	21:40
Air Blank		0.000	21:40
Diagnostics Check		OK	21:40

Data Download Date/Time: 02/16/2016 17:08

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARD,BENJAMIN,D
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: WARD,BENJAMIN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 01/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 19:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:56

SUBJECT'S LAST NAME: OSBOURN

DATE/TIME OF TEST: 01/29/2016 20:23:28

Diagnostics Check		OK	20:16
Air Blank		0.000	20:17
Control Test		0.079	20:17
Air Blank		0.000	20:18
Subject Sample	# 1	0.235	20:18
Breath Volume		3.085	20:18
Air Blank		0.000	20:19
Air Blank		0.000	20:21
Subject Sample	# 2	0.238	20:21
Breath Volume		3.921	20:21
Air Blank		0.000	20:22
Control Test		0.079	20:22
Air Blank		0.000	20:23
Diagnostics Check		OK	20:23

Data Download Date/Time: 02/16/2016 17:08

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPEAKMAN,ROBERT,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: SPEAKMAN,ROBERT
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 01/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:58

SUBJECT'S LAST NAME: SYPHER

DATE/TIME OF TEST: 01/30/2016 01:28:27

Diagnostics Check		OK	01:25	
Air Blank		0.000	01:25	
Control Test		0.080	01:26	
Air Blank		0.000	01:26	
Subject Sample	# 1		01:27	Subject Test Refused
Breath Volume		0.000	01:27	
Air Blank		0.000	01:27	
Control Test		0.080	01:27	
Air Blank		0.000	01:28	
Diagnostics Check		OK	01:28	

Data Download Date/Time: 02/16/2016 17:08

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREIRA,JOSE,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: PEREIRA,JOSE
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 01/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:49

SUBJECT'S LAST NAME: CAPECE

DATE/TIME OF TEST: 01/31/2016 01:18:59

Diagnostics Check		OK	01:10	
Air Blank		0.000	01:10	
Control Test		0.080	01:11	
Air Blank		0.000	01:11	
Subject Sample	# 1		01:14	No Sample Provided
Breath Volume		0.000	01:14	
Air Blank		0.000	01:15	
Air Blank		0.000	01:17	
Subject Sample	# 2		01:17	Subject Test Refused
Breath Volume		0.000	01:17	
Air Blank		0.000	01:18	
Control Test		0.079	01:18	
Air Blank		0.000	01:18	
Diagnostics Check		OK	01:18	

Data Download Date/Time: 02/16/2016 17:08

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREIRA,JOSE,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: PEREIRA,JOSE
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 01/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:53

SUBJECT'S LAST NAME: BORDELON

DATE/TIME OF TEST: 01/31/2016 04:27:20

Diagnostics Check		OK	04:15	
Air Blank		0.000	04:16	
Control Test		0.079	04:16	
Air Blank		0.000	04:17	
Subject Sample	# 1		04:20	No Sample Provided
Breath Volume		0.000	04:20	
Air Blank		0.000	04:20	
Air Blank		0.000	04:22	
Subject Sample	# 2		04:25	No Sample Provided
Breath Volume		0.000	04:25	
Air Blank		0.000	04:26	
Control Test		0.079	04:26	
Air Blank		0.000	04:27	
Diagnostics Check		OK	04:27	

Data Download Date/Time: 02/16/2016 17:08

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREIRA,JOSE,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: PEREIRA,JOSE
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 01/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:53

SUBJECT'S LAST NAME: BORDELON

DATE/TIME OF TEST: 01/31/2016 04:33:32

Diagnostics Check		OK	04:30	
Air Blank		0.000	04:30	
Control Test		0.078	04:31	
Air Blank		0.000	04:31	
Subject Sample	# 1		04:32	Subject Test Refused
Breath Volume		0.000	04:32	
Air Blank		0.000	04:32	
Control Test		0.079	04:32	
Air Blank		0.000	04:33	
Diagnostics Check		OK	04:33	

Data Download Date/Time: 02/16/2016 17:08

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005058

INSTRUMENT REGISTERED TO: CRESTVIEW PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARMER,GLENCHARLES,J
BREATH TEST OPERATOR AGENCY: Crestview PD
ARRESTING OFFICER: SEALS,JOSEPH
ARRESTING OFFICER AGENCY: Crestview PD
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09014080A1
DRY GAS EXPIRY DATE: 05/01/2016
VIOLATION CODE: DUI
ARREST TIME: 01:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: TRENT

DATE/TIME OF TEST: 01/01/2016 01:38:52

Diagnostics Check		OK	01:36	
Air Blank		0.000	01:36	
Control Test		0.080	01:36	
Air Blank		0.000	01:37	
Subject Sample	# 1		01:37	Subject Test Refused
Breath Volume		0.000	01:37	
Air Blank		0.000	01:37	
Control Test		0.080	01:38	
Air Blank		0.000	01:38	
Diagnostics Check		OK	01:38	

Data Download Date/Time: 01/04/2016 15:00

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005058

INSTRUMENT REGISTERED TO: CRESTVIEW PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SEALS,J,J
BREATH TEST OPERATOR AGENCY: Crestview PD
ARRESTING OFFICER: SEALS,J
ARRESTING OFFICER AGENCY: Crestview PD
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09014080A1
DRY GAS EXPIRY DATE: 05/01/2016
VIOLATION CODE: DUI
ARREST TIME: 20:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:46

SUBJECT'S LAST NAME: BENITES-PATINO

DATE/TIME OF TEST: 01/01/2016 21:17:15

Diagnostics Check		OK	21:10
Air Blank		0.000	21:10
Control Test		0.081	21:11
Air Blank		0.000	21:11
Subject Sample	# 1	0.171	21:12
Breath Volume		2.867	21:12
Air Blank		0.000	21:13
Air Blank		0.000	21:15
Subject Sample	# 2	0.168	21:15
Breath Volume		2.316	21:15
Air Blank		0.000	21:16
Control Test		0.079	21:16
Air Blank		0.000	21:17
Diagnostics Check		OK	21:17

Data Download Date/Time: 01/04/2016 15:00

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005058

INSTRUMENT REGISTERED TO: CRESTVIEW PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARMER,GLENCHARLES,J
BREATH TEST OPERATOR AGENCY: Crestview PD
ARRESTING OFFICER: FARMER,GLENCHARLES
ARRESTING OFFICER AGENCY: Crestview PD
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09014080A1
DRY GAS EXPIRY DATE: 05/01/2016
VIOLATION CODE: DUI
ARREST TIME: 23:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: BENNETT

DATE/TIME OF TEST: 01/16/2016 23:42:28

Diagnostics Check		OK	23:35
Air Blank		0.000	23:36
Control Test		0.080	23:36
Air Blank		0.000	23:37
Subject Sample	# 1	0.238	23:37
Breath Volume		3.066	23:37
Air Blank		0.000	23:38
Air Blank		0.000	23:40
Subject Sample	# 2	0.227	23:40
Breath Volume		1.925	23:40
Air Blank		0.000	23:41
Control Test		0.079	23:41
Air Blank		0.000	23:42
Diagnostics Check		OK	23:42

Data Download Date/Time: 01/28/2016 16:06

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005058

INSTRUMENT REGISTERED TO: CRESTVIEW PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARMER,GLENCHARLES,J
BREATH TEST OPERATOR AGENCY: Crestview PD
ARRESTING OFFICER: FARMER,GLENCHARLES
ARRESTING OFFICER AGENCY: Crestview PD
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09014080A1
DRY GAS EXPIRY DATE: 05/01/2016
VIOLATION CODE: DUI
ARREST TIME: 02:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: CREWS

DATE/TIME OF TEST: 01/23/2016 02:49:11

Diagnostics Check		OK	02:46	
Air Blank		0.000	02:46	
Control Test		0.081	02:47	
Air Blank		0.000	02:47	
Subject Sample	# 1		02:47	Subject Test Refused
Breath Volume		0.000	02:47	
Air Blank		0.000	02:48	
Control Test		0.081	02:48	
Air Blank		0.000	02:49	
Diagnostics Check		OK	02:49	

Data Download Date/Time: 01/28/2016 16:06

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005058

INSTRUMENT REGISTERED TO: CRESTVIEW PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARMER,GLENCHARLES,J
BREATH TEST OPERATOR AGENCY: Crestview PD
ARRESTING OFFICER: FARMER,GLENCHARLES
ARRESTING OFFICER AGENCY: Crestview PD
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09014080A1
DRY GAS EXPIRY DATE: 05/01/2016
VIOLATION CODE: DUI
ARREST TIME: 18:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:22

SUBJECT'S LAST NAME: ALLDREDGE

DATE/TIME OF TEST: 01/24/2016 18:49:10

Diagnostics Check		OK	18:43
Air Blank		0.000	18:43
Control Test		0.080	18:43
Air Blank		0.000	18:44
Subject Sample	# 1	0.164	18:44
Breath Volume		1.375	18:44
Air Blank		0.000	18:45
Air Blank		0.000	18:47
Subject Sample	# 2	0.158	18:47
Breath Volume		2.664	18:47
Air Blank		0.000	18:48
Control Test		0.080	18:48
Air Blank		0.000	18:49
Diagnostics Check		OK	18:49

Data Download Date/Time: 01/28/2016 16:06

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005058

INSTRUMENT REGISTERED TO: CRESTVIEW PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOKEN,JEFFREY,W
BREATH TEST OPERATOR AGENCY: Crestview PD
ARRESTING OFFICER: STANLEY,DAVID
ARRESTING OFFICER AGENCY: Crestview PD
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09014080A1
DRY GAS EXPIRY DATE: 05/01/2016
VIOLATION CODE: DUI
ARREST TIME: 16:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:06

SUBJECT'S LAST NAME: MATHIS

DATE/TIME OF TEST: 01/25/2016 17:36:07

Diagnostics Check		OK	17:26
Air Blank		0.000	17:26
Control Test		0.080	17:27
Air Blank		0.000	17:27
Subject Sample	# 1	0.215	17:31
Breath Volume		2.484	17:31
Air Blank		0.000	17:31
Air Blank		0.000	17:33
Subject Sample	# 2	0.214	17:34
Breath Volume		1.226	17:34
Air Blank		0.000	17:35
Control Test		0.079	17:35
Air Blank		0.000	17:36
Diagnostics Check		OK	17:36

Data Download Date/Time: 01/28/2016 16:06

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005076

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FOWLER,TYLER,J
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: WEAVER,RICHARD
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 12/28/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28814080A3
DRY GAS EXPIRY DATE: 11/05/2016
VIOLATION CODE: DUI
ARREST TIME: 00:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:58

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 01/01/2016 00:29:09

Diagnostics Check		OK	00:22
Air Blank		0.000	00:23
Control Test		0.077	00:23
Air Blank		0.000	00:24
Subject Sample	# 1	0.163	00:24
Breath Volume		2.035	00:24
Air Blank		0.000	00:25
Air Blank		0.000	00:27
Subject Sample	# 2	0.159	00:27
Breath Volume		2.210	00:27
Air Blank		0.000	00:28
Control Test		0.077	00:28
Air Blank		0.000	00:29
Diagnostics Check		OK	00:29

Data Download Date/Time: 01/27/2016 09:14

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005076

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TUCK,DEPUTY,B
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: TUCK,DEPUTY B
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 12/28/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28814080A3
DRY GAS EXPIRY DATE: 11/05/2016
VIOLATION CODE: DUI
ARREST TIME: 00:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: MOORE

DATE/TIME OF TEST: 01/01/2016 00:47:25

Diagnostics Check		OK	00:40
Air Blank		0.000	00:41
Control Test		0.077	00:41
Air Blank		0.000	00:42
Subject Sample	# 1	0.151	00:43
Breath Volume		4.097	00:43
Air Blank		0.000	00:43
Air Blank		0.000	00:45
Subject Sample	# 2	0.134	00:45
Breath Volume		1.718	00:45
Air Blank		0.000	00:46
Control Test		0.077	00:46
Air Blank		0.000	00:47
Diagnostics Check		OK	00:47

Data Download Date/Time: 01/27/2016 09:14

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005076

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FOWLER,TYLER,J
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: FOWLER,TYLER
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 12/28/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28814080A3
DRY GAS EXPIRY DATE: 11/05/2016
VIOLATION CODE: DUI
ARREST TIME: 00:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: BOYD

DATE/TIME OF TEST: 01/02/2016 00:22:50

Diagnostics Check		OK	00:16
Air Blank		0.000	00:16
Control Test		0.079	00:17
Air Blank		0.000	00:17
Subject Sample	# 1	0.230	00:18
Breath Volume		2.199	00:18
Air Blank		0.000	00:19
Air Blank		0.000	00:20
Subject Sample	# 2	0.228	00:21
Breath Volume		1.316	00:21
Air Blank		0.000	00:21
Control Test		0.077	00:22
Air Blank		0.000	00:22
Diagnostics Check		OK	00:22

Data Download Date/Time: 01/27/2016 09:14

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005076

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLINT,SERGEANT,C
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: FLINT,SERGEANT
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 12/28/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28814080A3
DRY GAS EXPIRY DATE: 11/05/2016
VIOLATION CODE: DUI
ARREST TIME: 22:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:33

SUBJECT'S LAST NAME: MATHEWS

DATE/TIME OF TEST: 01/08/2016 23:05:38

Diagnostics Check		OK	22:59
Air Blank		0.000	22:59
Control Test		0.078	23:00
Air Blank		0.000	23:00
Subject Sample	# 1	0.151	23:01
Breath Volume		1.457	23:01
Air Blank		0.000	23:01
Air Blank		0.000	23:03
Subject Sample	# 2	0.157	23:04
Breath Volume		1.632	23:04
Air Blank		0.000	23:04
Control Test		0.078	23:05
Air Blank		0.000	23:05
Diagnostics Check		OK	23:05

Data Download Date/Time: 01/27/2016 09:14

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005076

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FOWLER,TYLER,J
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: FOWLER,TYLER
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 01/27/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28814080A3
DRY GAS EXPIRY DATE: 11/05/2016
VIOLATION CODE: DUI
ARREST TIME: 01:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:36

SUBJECT'S LAST NAME: ORTIZ PONCE

DATE/TIME OF TEST: 01/31/2016 02:10:06

Diagnostics Check		OK	02:00	
Air Blank		0.000	02:00	
Control Test		0.079	02:01	
Air Blank		0.000	02:01	
Subject Sample	# 1	0.189	02:04	
Breath Volume		3.265	02:04	
Air Blank		0.000	02:05	
Air Blank		0.000	02:07	
Subject Sample	# 2	0.175	02:08	
Breath Volume		2.378	02:08	
Air Blank		0.000	02:09	
Control Test		0.074	02:09	Control Outside Tolerance
Air Blank		0.000	02:10	

Data Download Date/Time: 02/23/2016 09:01

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005076

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FOWLER,TYLER,J
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: FOWLER,TYLER
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 01/27/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28814080A3
DRY GAS EXPIRY DATE: 11/05/2016
VIOLATION CODE: DUI
ARREST TIME: 01:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:16

SUBJECT'S LAST NAME: ORTIZ PONCE

DATE/TIME OF TEST: 01/31/2016 02:44:06

Diagnostics Check		OK	02:37
Air Blank		0.000	02:38
Control Test		0.078	02:38
Air Blank		0.000	02:38
Subject Sample	# 1	0.173	02:39
Breath Volume		1.503	02:39
Air Blank		0.000	02:40
Air Blank		0.000	02:42
Subject Sample	# 2	0.192	02:42
Breath Volume		2.824	02:42
Air Blank		0.000	02:43
Control Test		0.078	02:43
Air Blank		0.000	02:44
Diagnostics Check		OK	02:44

Data Download Date/Time: 02/23/2016 09:01

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005077

INSTRUMENT REGISTERED TO: NORTH MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PINEDA,MIGUEL,A
BREATH TEST OPERATOR AGENCY: North Miami PD
ARRESTING OFFICER: ALEDDA,JONATHAN
ARRESTING OFFICER AGENCY: North Miami PD
LAST AGENCY INSPECTION: 01/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06415080A2
DRY GAS EXPIRY DATE: 04/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: ROMAN

DATE/TIME OF TEST: 01/19/2016 23:54:48

Diagnostics Check		OK	23:46
Air Blank		0.000	23:47
Control Test		0.079	23:47
Air Blank		0.000	23:48
Subject Sample	# 1	0.245	23:49
Breath Volume		1.191	23:49
Air Blank		0.000	23:49
Air Blank		0.000	23:51
Subject Sample	# 2	0.253	23:53
Breath Volume		1.656	23:53
Air Blank		0.000	23:53
Control Test		0.078	23:54
Air Blank		0.000	23:54
Diagnostics Check		OK	23:54

Data Download Date/Time: 02/18/2016 15:39

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY,PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ANGELICCHI,J
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/10/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:18

SUBJECT'S LAST NAME: REYTOR

DATE/TIME OF TEST: 01/01/2016 02:45:24

Diagnostics Check		OK	02:38
Air Blank		0.000	02:39
Control Test		0.076	02:39
Air Blank		0.000	02:40
Subject Sample	# 1	0.149	02:40
Breath Volume		1.996	02:40
Air Blank		0.000	02:41
Air Blank		0.000	02:43
Subject Sample	# 2	0.151	02:43
Breath Volume		1.988	02:43
Air Blank		0.000	02:44
Control Test		0.076	02:44
Air Blank		0.000	02:45
Diagnostics Check		OK	02:45

Data Download Date/Time: 01/04/2016 19:11

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRABAND,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 19:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:21

SUBJECT'S LAST NAME: YOUNCE

DATE/TIME OF TEST: 01/04/2016 20:49:45

Diagnostics Check		OK	20:42
Air Blank		0.000	20:42
Control Test		0.078	20:43
Air Blank		0.000	20:43
Subject Sample	# 1	0.000	20:45
Breath Volume		1.179	20:45
Air Blank		0.000	20:45
Air Blank		0.000	20:47
Subject Sample	# 2	0.000	20:48
Breath Volume		1.285	20:48
Air Blank		0.000	20:48
Control Test		0.080	20:49
Air Blank		0.000	20:49
Diagnostics Check		OK	20:49

Data Download Date/Time: 02/01/2016 20:53

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:18

SUBJECT'S LAST NAME: ARROYAVE CABRERA

DATE/TIME OF TEST: 01/06/2016 02:46:38

Diagnostics Check		OK	02:40
Air Blank		0.000	02:40
Control Test		0.079	02:40
Air Blank		0.000	02:41
Subject Sample	# 1	0.087	02:42
Breath Volume		3.167	02:42
Air Blank		0.000	02:42
Air Blank		0.000	02:44
Subject Sample	# 2	0.085	02:45
Breath Volume		2.492	02:45
Air Blank		0.000	02:45
Control Test		0.081	02:46
Air Blank		0.000	02:46
Diagnostics Check		OK	02:46

Data Download Date/Time: 02/01/2016 20:53

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VAN-TREESE,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: PRISCO

DATE/TIME OF TEST: 01/07/2016 00:03:26

Diagnostics Check		OK	23:57
Air Blank		0.000	23:57
Control Test		0.079	23:57
Air Blank		0.000	23:58
Subject Sample	# 2	0.000	23:59
Breath Volume		1.343	23:59
Air Blank		0.000	23:59
Air Blank		0.000	00:01
Subject Sample	# 1	0.000	00:02
Breath Volume		1.664	00:02
Air Blank		0.000	00:02
Control Test		0.080	00:02
Air Blank		0.000	00:03
Diagnostics Check		OK	00:03

Data Download Date/Time: 02/01/2016 20:53

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WOODEN,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:51

SUBJECT'S LAST NAME: WARDROP

DATE/TIME OF TEST: 01/08/2016 02:21:31

Diagnostics Check		OK	02:12
Air Blank		0.000	02:13
Control Test		0.078	02:13
Air Blank		0.000	02:14
Subject Sample	# 1	0.111	02:16
Breath Volume		1.566	02:16
Air Blank		0.000	02:17
Air Blank		0.000	02:19
Subject Sample	# 2	0.114	02:19
Breath Volume		1.640	02:19
Air Blank		0.000	02:20
Control Test		0.081	02:20
Air Blank		0.000	02:21
Diagnostics Check		OK	02:21

Data Download Date/Time: 02/01/2016 20:53

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HERNANDEZ,A
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:12

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 01/09/2016 02:39:00

Diagnostics Check		OK	02:32
Air Blank		0.000	02:33
Control Test		0.078	02:33
Air Blank		0.000	02:33
Subject Sample	# 1	0.125	02:34
Breath Volume		2.558	02:34
Air Blank		0.000	02:35
Air Blank		0.000	02:36
Subject Sample	# 2	0.126	02:37
Breath Volume		1.203	02:37
Air Blank		0.000	02:38
Control Test		0.079	02:38
Air Blank		0.000	02:38
Diagnostics Check		OK	02:39

Data Download Date/Time: 02/01/2016 20:53

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FRANKLIN,R
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:53

SUBJECT'S LAST NAME: SPECTOR

DATE/TIME OF TEST: 01/09/2016 03:26:35

Diagnostics Check		OK	03:14	
Air Blank		0.000	03:14	
Control Test		0.078	03:14	
Air Blank		0.000	03:15	
Subject Sample	# 1	0.160	03:18	Volume Not Met
Breath Volume		0.136	03:18	
Air Blank		0.000	03:19	
Air Blank		0.000	03:20	
Subject Sample	# 2	0.207	03:21	No .020 agreement
Breath Volume		2.550	03:21	
Air Blank		0.000	03:22	
Air Blank		0.000	03:24	
Subject Sample	# 3	0.184	03:24	No .020 agreement
Breath Volume		1.335	03:24	
Air Blank		0.000	03:25	
Control Test		0.079	03:25	
Air Blank		0.000	03:26	
Diagnostics Check		OK	03:26	

Data Download Date/Time: 02/01/2016 20:53

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FRANKLIN,R
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:53

SUBJECT'S LAST NAME: SPECTOR

DATE/TIME OF TEST: 01/09/2016 03:51:16

Diagnostics Check		OK	03:35
Air Blank		0.000	03:36
Control Test		0.078	03:36
Air Blank		0.000	03:36
Subject Sample	# 1	0.200	03:39
Breath Volume		1.660	03:39
Air Blank		0.000	03:40
Air Blank		0.000	03:42
Subject Sample	# 2	0.171	03:43
Breath Volume		2.003	03:43
Air Blank		0.000	03:44
Air Blank		0.000	03:46
Subject Sample	# 3	0.170	03:49
Breath Volume		1.468	03:49
Air Blank		0.000	03:50
Control Test		0.075	03:50
Air Blank		0.000	03:51
Diagnostics Check		OK	03:51

Data Download Date/Time: 02/01/2016 20:53

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK,CASEY,S
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HALL,K
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: VAZQUEZ CRUZ

DATE/TIME OF TEST: 01/09/2016 04:47:38

Diagnostics Check		OK	04:40
Air Blank		0.000	04:41
Control Test		0.075	04:41
Air Blank		0.000	04:42
Subject Sample	# 1	0.169	04:43
Breath Volume		1.347	04:43
Air Blank		0.000	04:43
Air Blank		0.000	04:45
Subject Sample	# 2	0.164	04:46
Breath Volume		2.660	04:46
Air Blank		0.000	04:46
Control Test		0.076	04:47
Air Blank		0.000	04:47
Diagnostics Check		OK	04:47

Data Download Date/Time: 02/01/2016 20:53

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMMONS,ANGELA,V
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WARD,E
ARRESTING OFFICER AGENCY: Plant City PD
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 17:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:02

SUBJECT'S LAST NAME: HARPER JR

DATE/TIME OF TEST: 01/09/2016 18:30:39

Diagnostics Check		OK	18:24
Air Blank		0.000	18:24
Control Test		0.078	18:24
Air Blank		0.000	18:25
Subject Sample	# 1	0.240	18:26
Breath Volume		3.121	18:26
Air Blank		0.000	18:26
Air Blank		0.000	18:28
Subject Sample	# 2	0.236	18:29
Breath Volume		3.183	18:29
Air Blank		0.000	18:29
Control Test		0.079	18:30
Air Blank		0.000	18:30
Diagnostics Check		OK	18:30

Data Download Date/Time: 02/01/2016 20:53

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:32

SUBJECT'S LAST NAME: AUSTIN

DATE/TIME OF TEST: 01/10/2016 01:01:09

Diagnostics Check		OK	00:53
Air Blank		0.000	00:53
Control Test		0.078	00:54
Air Blank		0.000	00:54
Subject Sample	# 1	0.135	00:56
Breath Volume		1.328	00:56
Air Blank		0.000	00:56
Air Blank		0.000	00:58
Subject Sample	# 2	0.135	00:59
Breath Volume		1.328	00:59
Air Blank		0.000	01:00
Control Test		0.081	01:00
Air Blank		0.000	01:01
Diagnostics Check		OK	01:01

Data Download Date/Time: 02/01/2016 20:53

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY,PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FRANKLIN,R
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: HUYNH

DATE/TIME OF TEST: 01/10/2016 02:32:40

Diagnostics Check		OK	02:26
Air Blank		0.000	02:26
Control Test		0.078	02:27
Air Blank		0.000	02:27
Subject Sample	# 1	0.123	02:28
Breath Volume		1.597	02:28
Air Blank		0.000	02:28
Air Blank		0.000	02:30
Subject Sample	# 2	0.120	02:31
Breath Volume		2.476	02:31
Air Blank		0.000	02:31
Control Test		0.078	02:32
Air Blank		0.000	02:32
Diagnostics Check		OK	02:32

Data Download Date/Time: 02/01/2016 20:53

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMMONS,ANGELA,V
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: THORNBERRY,J
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 17:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:50

SUBJECT'S LAST NAME: COOPER

DATE/TIME OF TEST: 01/10/2016 18:15:09

Diagnostics Check		OK	18:11	
Air Blank		0.000	18:11	
Control Test		0.078	18:12	
Air Blank		0.000	18:12	
Subject Sample	# 1	0.189	18:13	
Breath Volume		4.496	18:13	
Air Blank			18:13	RFI Detect
Air Blank		0.000	18:14	
Control Test		0.079	18:14	
Air Blank		0.000	18:15	
Diagnostics Check		OK	18:15	

Data Download Date/Time: 02/01/2016 20:53

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMMONS,ANGELA,V
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: THORNBERRY,J
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 17:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:50

SUBJECT'S LAST NAME: COOPER

DATE/TIME OF TEST: 01/10/2016 18:24:55

Diagnostics Check		OK	18:17
Air Blank		0.000	18:17
Control Test		0.077	18:17
Air Blank		0.000	18:18
Subject Sample	# 1	0.189	18:20
Breath Volume		4.480	18:20
Air Blank		0.000	18:20
Air Blank		0.000	18:22
Subject Sample	# 2	0.191	18:23
Breath Volume		4.613	18:23
Air Blank		0.000	18:24
Control Test		0.078	18:24
Air Blank		0.000	18:24
Diagnostics Check		OK	18:24

Data Download Date/Time: 02/01/2016 20:53

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:27

SUBJECT'S LAST NAME: GAINES

DATE/TIME OF TEST: 01/10/2016 21:54:15

Diagnostics Check		OK	21:47
Air Blank		0.000	21:48
Control Test		0.078	21:48
Air Blank		0.000	21:48
Subject Sample	# 1	0.143	21:49
Breath Volume		2.335	21:49
Air Blank		0.000	21:50
Air Blank		0.000	21:52
Subject Sample	# 2	0.137	21:52
Breath Volume		1.562	21:52
Air Blank		0.000	21:53
Control Test		0.078	21:53
Air Blank		0.000	21:54
Diagnostics Check		OK	21:54

Data Download Date/Time: 02/01/2016 20:53

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SARRO,TAYLOR,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GOFF,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:21

SUBJECT'S LAST NAME: DWYER

DATE/TIME OF TEST: 01/11/2016 03:48:21

Diagnostics Check		OK	03:42
Air Blank		0.000	03:42
Control Test		0.079	03:42
Air Blank		0.000	03:43
Subject Sample	# 1	0.154	03:43
Breath Volume		1.554	03:43
Air Blank		0.000	03:44
Air Blank		0.000	03:46
Subject Sample	# 2	0.159	03:46
Breath Volume		1.183	03:46
Air Blank		0.000	03:47
Control Test		0.078	03:47
Air Blank		0.000	03:48
Diagnostics Check		OK	03:48

Data Download Date/Time: 02/01/2016 20:53

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY,PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GONZALEZ,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 19:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:12

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 01/11/2016 20:39:38

Diagnostics Check		OK	20:32
Air Blank		0.000	20:32
Control Test		0.079	20:32
Air Blank		0.000	20:33
Subject Sample	# 1	0.154	20:34
Breath Volume		1.171	20:34
Air Blank		0.000	20:34
Air Blank		0.000	20:36
Subject Sample	# 2	0.147	20:37
Breath Volume		1.355	20:37
Air Blank		0.000	20:38
Control Test		0.080	20:38
Air Blank		0.000	20:39
Diagnostics Check		OK	20:39

Data Download Date/Time: 02/01/2016 20:53

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNYDER,MELONIE,E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ANDERSON,C
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 06:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:06

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 01/12/2016 07:33:27

Diagnostics Check		OK	07:26
Air Blank		0.000	07:26
Control Test		0.079	07:27
Air Blank		0.000	07:27
Subject Sample	# 1	0.148	07:28
Breath Volume		1.187	07:28
Air Blank		0.000	07:29
Air Blank		0.000	07:31
Subject Sample	# 2	0.147	07:31
Breath Volume		1.683	07:31
Air Blank		0.000	07:32
Control Test		0.078	07:32
Air Blank		0.000	07:33
Diagnostics Check		OK	07:33

Data Download Date/Time: 02/01/2016 20:53

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRABAND,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:51

SUBJECT'S LAST NAME: CONRAD III

DATE/TIME OF TEST: 01/13/2016 03:26:05

Diagnostics Check		OK	03:12	
Air Blank		0.000	03:12	
Control Test		0.079	03:13	
Air Blank		0.000	03:13	
Subject Sample	# 1		03:16	No Sample Provided
Breath Volume		0.000	03:16	
Air Blank		0.000	03:17	
Air Blank		0.000	03:19	
Subject Sample	# 2	0.097	03:20	
Breath Volume		2.585	03:20	
Air Blank		0.000	03:21	
Air Blank		0.000	03:23	
Subject Sample	# 3	0.100	03:24	
Breath Volume		1.531	03:24	
Air Blank		0.000	03:25	
Control Test		0.080	03:25	
Air Blank		0.000	03:26	
Diagnostics Check		OK	03:26	

Data Download Date/Time: 02/01/2016 20:53

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FRANKLIN,R
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:46

SUBJECT'S LAST NAME: RAMOS

DATE/TIME OF TEST: 01/14/2016 00:14:54

Diagnostics Check		OK	00:06
Air Blank		0.000	00:06
Control Test		0.079	00:07
Air Blank		0.000	00:07
Subject Sample	# 1	0.120	00:09
Breath Volume		1.300	00:09
Air Blank		0.000	00:10
Air Blank		0.000	00:12
Subject Sample	# 2	0.117	00:13
Breath Volume		1.640	00:13
Air Blank		0.000	00:13
Control Test		0.079	00:14
Air Blank		0.000	00:14
Diagnostics Check		OK	00:14

Data Download Date/Time: 02/01/2016 20:53

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY,PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FRANKLIN,R
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:45

SUBJECT'S LAST NAME: WRIGHT

DATE/TIME OF TEST: 01/15/2016 05:09:22

Diagnostics Check		OK	05:05	
Air Blank		0.000	05:05	
Control Test		0.079	05:06	
Air Blank		0.000	05:06	
Subject Sample	# 1		05:07	Subject Test Refused
Breath Volume		0.000	05:07	
Air Blank		0.000	05:08	
Control Test		0.078	05:08	
Air Blank		0.000	05:09	
Diagnostics Check		OK	05:09	

Data Download Date/Time: 02/01/2016 20:53

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARR,BARBARA,H
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RUSSELL,K
ARRESTING OFFICER AGENCY: Plant City PD
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 14:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:30

SUBJECT'S LAST NAME: TESEDA OBANDO

DATE/TIME OF TEST: 01/15/2016 16:00:28

Diagnostics Check		OK	15:54
Air Blank		0.000	15:54
Control Test		0.078	15:55
Air Blank		0.000	15:55
Subject Sample	# 1	0.147	15:56
Breath Volume		1.296	15:56
Air Blank		0.000	15:56
Air Blank		0.000	15:58
Subject Sample	# 2	0.155	15:58
Breath Volume		2.011	15:58
Air Blank		0.000	15:59
Control Test		0.078	15:59
Air Blank		0.000	16:00
Diagnostics Check		OK	16:00

Data Download Date/Time: 02/01/2016 20:53

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SARRO,TAYLOR,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:43

SUBJECT'S LAST NAME: MCCLURKIN

DATE/TIME OF TEST: 01/16/2016 00:10:24

Diagnostics Check		OK	00:03
Air Blank		0.000	00:04
Control Test		0.079	00:04
Air Blank		0.000	00:05
Subject Sample	# 1	0.125	00:05
Breath Volume		1.257	00:05
Air Blank		0.000	00:06
Air Blank		0.000	00:08
Subject Sample	# 2	0.122	00:08
Breath Volume		1.281	00:08
Air Blank		0.000	00:09
Control Test		0.078	00:09
Air Blank		0.000	00:10
Diagnostics Check		OK	00:10

Data Download Date/Time: 02/01/2016 20:53

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MULLEN,E
ARRESTING OFFICER AGENCY: Seminole Res (Hills)
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:56

SUBJECT'S LAST NAME: BARRETT

DATE/TIME OF TEST: 01/16/2016 02:23:14

Diagnostics Check		OK	02:20	
Air Blank		0.000	02:20	
Control Test		0.079	02:20	
Air Blank		0.000	02:21	
Subject Sample	# 1		02:21	Subject Test Refused
Breath Volume		0.000	02:21	
Air Blank		0.000	02:22	
Control Test		0.078	02:22	
Air Blank		0.000	02:23	
Diagnostics Check		OK	02:23	

Data Download Date/Time: 02/01/2016 20:53

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY,PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HBSON,D
ARRESTING OFFICER AGENCY: U of So FL Campus PD
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: MARTIN

DATE/TIME OF TEST: 01/16/2016 03:26:24

Diagnostics Check		OK	03:20
Air Blank		0.000	03:20
Control Test		0.078	03:20
Air Blank		0.000	03:21
Subject Sample	# 1	0.117	03:22
Breath Volume		2.207	03:22
Air Blank		0.000	03:22
Air Blank		0.000	03:24
Subject Sample	# 2	0.113	03:24
Breath Volume		1.609	03:24
Air Blank		0.000	03:25
Control Test		0.078	03:25
Air Blank		0.000	03:26
Diagnostics Check		OK	03:26

Data Download Date/Time: 02/01/2016 20:53

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SARRO,TAYLOR,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DE JESUS-COLON,S
ARRESTING OFFICER AGENCY: Temple Terrace PD
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:14

SUBJECT'S LAST NAME: BARKLEY

DATE/TIME OF TEST: 01/16/2016 03:50:59

Diagnostics Check		OK	03:37	
Air Blank		0.000	03:37	
Control Test		0.078	03:38	
Air Blank		0.000	03:38	
Subject Sample	# 1	0.135	03:40	
Breath Volume		1.242	03:40	
Air Blank		0.000	03:41	
Air Blank		0.000	03:43	
Subject Sample	# 2	0.125	03:46	Volume Not Met
Breath Volume		0.992	03:46	
Air Blank		0.000	03:47	
Air Blank		0.000	03:49	
Subject Sample	# 3	0.134	03:49	
Breath Volume		1.539	03:49	
Air Blank		0.000	03:50	
Control Test		0.079	03:50	
Air Blank		0.000	03:50	
Diagnostics Check		OK	03:50	

Data Download Date/Time: 02/01/2016 20:53

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNYDER,MELONIE,E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MCCRANIE,B
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 12:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:52

SUBJECT'S LAST NAME: MURRAY

DATE/TIME OF TEST: 01/16/2016 14:16:37

Diagnostics Check		OK	14:14	
Air Blank		0.000	14:15	
Control Test		0.078	14:15	
Air Blank		0.000	14:15	
Subject Sample	# 1		14:16	Sequence Aborted
Breath Volume		0.000	14:16	
Air Blank		0.000	14:16	

Data Download Date/Time: 02/01/2016 20:53

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNYDER,MELONIE,E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MCCRANIE,B
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 12:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:17

SUBJECT'S LAST NAME: MURRAY

DATE/TIME OF TEST: 01/16/2016 14:44:06

Diagnostics Check		OK	14:37
Air Blank		0.000	14:37
Control Test		0.078	14:38
Air Blank		0.000	14:38
Subject Sample	# 1	0.188	14:39
Breath Volume		2.972	14:39
Air Blank		0.000	14:40
Air Blank		0.000	14:41
Subject Sample	# 2	0.194	14:42
Breath Volume		2.402	14:42
Air Blank		0.000	14:43
Control Test		0.078	14:43
Air Blank		0.000	14:44
Diagnostics Check		OK	14:44

Data Download Date/Time: 02/01/2016 20:53

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNYDER,MELONIE,E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: CORTES,P
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 15:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:11

SUBJECT'S LAST NAME: RUKSE

DATE/TIME OF TEST: 01/16/2016 16:39:19

Diagnostics Check		OK	16:32
Air Blank		0.000	16:33
Control Test		0.079	16:33
Air Blank		0.000	16:34
Subject Sample	# 1	0.308	16:34
Breath Volume		1.941	16:34
Air Blank		0.000	16:35
Air Blank		0.000	16:37
Subject Sample	# 2	0.307	16:37
Breath Volume		1.703	16:37
Air Blank		0.000	16:38
Control Test		0.078	16:38
Air Blank		0.000	16:39
Diagnostics Check		OK	16:39

Data Download Date/Time: 02/01/2016 20:53

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:47

SUBJECT'S LAST NAME: MALAVE JR

DATE/TIME OF TEST: 01/16/2016 22:16:03

Diagnostics Check		OK	22:07
Air Blank		0.000	22:07
Control Test		0.078	22:08
Air Blank		0.000	22:08
Subject Sample	# 1	0.269	22:08
Breath Volume		1.535	22:08
Air Blank		0.000	22:09
Air Blank		0.000	22:11
Subject Sample	# 2	0.261	22:14
Breath Volume		1.597	22:14
Air Blank		0.000	22:15
Control Test		0.081	22:15
Air Blank		0.000	22:15
Diagnostics Check		OK	22:16

Data Download Date/Time: 02/01/2016 20:53

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK,CASEY,S
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SIMPKINS,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: NAVAS

DATE/TIME OF TEST: 01/16/2016 23:29:40

Diagnostics Check		OK	23:21
Air Blank		0.000	23:22
Control Test		0.078	23:22
Air Blank		0.000	23:22
Subject Sample	# 1	0.200	23:24
Breath Volume		2.015	23:24
Air Blank		0.000	23:25
Air Blank		0.000	23:27
Subject Sample	# 2	0.210	23:28
Breath Volume		2.066	23:28
Air Blank		0.000	23:28
Control Test		0.077	23:29
Air Blank		0.000	23:29
Diagnostics Check		OK	23:29

Data Download Date/Time: 02/01/2016 20:53

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BLANCHARD,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: BLUE

DATE/TIME OF TEST: 01/17/2016 00:38:30

Diagnostics Check		OK	00:35	
Air Blank		0.000	00:36	
Control Test		0.078	00:36	
Air Blank		0.000	00:36	
Subject Sample	# 1		00:37	Subject Test Refused
Breath Volume		0.000	00:37	
Air Blank		0.000	00:37	
Control Test		0.078	00:37	
Air Blank		0.000	00:38	
Diagnostics Check		OK	00:38	

Data Download Date/Time: 02/01/2016 20:53

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MILLER,J
ARRESTING OFFICER AGENCY: Temple Terrace PD
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:58

SUBJECT'S LAST NAME: PORTER

DATE/TIME OF TEST: 01/17/2016 21:24:59

Diagnostics Check		OK	21:18
Air Blank		0.000	21:18
Control Test		0.078	21:19
Air Blank		0.000	21:19
Subject Sample	# 1	0.126	21:20
Breath Volume		3.050	21:20
Air Blank		0.000	21:21
Air Blank		0.000	21:22
Subject Sample	# 2	0.125	21:23
Breath Volume		2.925	21:23
Air Blank		0.000	21:24
Control Test		0.078	21:24
Air Blank		0.000	21:24
Diagnostics Check		OK	21:24

Data Download Date/Time: 02/01/2016 20:53

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: FRIEDRICHS

DATE/TIME OF TEST: 01/18/2016 04:47:01

Diagnostics Check		OK	04:40
Air Blank		0.000	04:41
Control Test		0.079	04:41
Air Blank		0.000	04:41
Subject Sample	# 1	0.092	04:42
Breath Volume		1.937	04:42
Air Blank		0.000	04:43
Air Blank		0.000	04:44
Subject Sample	# 2	0.086	04:45
Breath Volume		1.312	04:45
Air Blank		0.000	04:46
Control Test		0.079	04:46
Air Blank		0.000	04:46
Diagnostics Check		OK	04:47

Data Download Date/Time: 02/01/2016 20:53

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNYDER,MELONIE,E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MATAS,T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 17:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:33

SUBJECT'S LAST NAME: CANTERBURY

DATE/TIME OF TEST: 01/18/2016 18:56:17

Diagnostics Check		OK	18:53	
Air Blank		0.000	18:53	
Control Test		0.079	18:54	
Air Blank		0.000	18:54	
Subject Sample	# 1		18:54	Subject Test Refused
Breath Volume		0.000	18:54	
Air Blank		0.000	18:55	
Control Test		0.079	18:55	
Air Blank		0.000	18:56	
Diagnostics Check		OK	18:56	

Data Download Date/Time: 02/01/2016 20:53

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK,CASEY,S
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ANGELICCHI,J
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:18

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 01/18/2016 21:46:29

Diagnostics Check		OK	21:39
Air Blank		0.000	21:39
Control Test		0.079	21:39
Air Blank		0.000	21:40
Subject Sample	# 1	0.257	21:41
Breath Volume		1.761	21:41
Air Blank		0.000	21:42
Air Blank		0.000	21:44
Subject Sample	# 2	0.270	21:44
Breath Volume		1.562	21:44
Air Blank		0.000	21:45
Control Test		0.079	21:45
Air Blank		0.000	21:46
Diagnostics Check		OK	21:46

Data Download Date/Time: 02/01/2016 20:53

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DE JESUS,V
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:59

SUBJECT'S LAST NAME: LILLARD

DATE/TIME OF TEST: 01/19/2016 05:26:01

Diagnostics Check		OK	05:19
Air Blank		0.000	05:19
Control Test		0.079	05:20
Air Blank		0.000	05:20
Subject Sample	# 1	0.145	05:21
Breath Volume		3.027	05:21
Air Blank		0.000	05:22
Air Blank		0.000	05:23
Subject Sample	# 2	0.150	05:24
Breath Volume		3.007	05:24
Air Blank		0.000	05:25
Control Test		0.079	05:25
Air Blank		0.000	05:25
Diagnostics Check		OK	05:26

Data Download Date/Time: 02/01/2016 20:53

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SARRO,TAYLOR,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARLAUG,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:51

SUBJECT'S LAST NAME: VICK

DATE/TIME OF TEST: 01/20/2016 01:17:40

Diagnostics Check		OK	01:11
Air Blank		0.000	01:11
Control Test		0.079	01:12
Air Blank		0.000	01:12
Subject Sample	# 1	0.157	01:13
Breath Volume		2.832	01:13
Air Blank		0.000	01:13
Air Blank		0.000	01:15
Subject Sample	# 2	0.156	01:16
Breath Volume		2.996	01:16
Air Blank		0.000	01:16
Control Test		0.079	01:17
Air Blank		0.000	01:17
Diagnostics Check		OK	01:17

Data Download Date/Time: 02/01/2016 20:53

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SARRO,TAYLOR,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SIMPKINS,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: AVILA

DATE/TIME OF TEST: 01/21/2016 00:02:03

Diagnostics Check		OK	23:56
Air Blank		0.000	23:56
Control Test		0.079	23:56
Air Blank		0.000	23:57
Subject Sample	# 2	0.000	23:57
Breath Volume		1.730	23:57
Air Blank		0.000	23:58
Air Blank		0.000	00:00
Subject Sample	# 1	0.000	00:00
Breath Volume		2.246	00:00
Air Blank		0.000	00:01
Control Test		0.079	00:01
Air Blank		0.000	00:01
Diagnostics Check		OK	00:02

Data Download Date/Time: 02/01/2016 20:53

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRABAND,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:54

SUBJECT'S LAST NAME: BAGLIN

DATE/TIME OF TEST: 01/21/2016 05:22:14

Diagnostics Check		OK	05:15
Air Blank		0.000	05:15
Control Test		0.079	05:16
Air Blank		0.000	05:16
Subject Sample	# 1	0.110	05:17
Breath Volume		2.941	05:17
Air Blank		0.000	05:18
Air Blank		0.000	05:20
Subject Sample	# 2	0.114	05:20
Breath Volume		2.941	05:20
Air Blank		0.000	05:21
Control Test		0.079	05:21
Air Blank		0.000	05:22
Diagnostics Check		OK	05:22

Data Download Date/Time: 02/01/2016 20:53

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VAN TREESE,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:56

SUBJECT'S LAST NAME: CASTILLO

DATE/TIME OF TEST: 01/22/2016 02:23:13

Diagnostics Check		OK	02:16
Air Blank		0.000	02:17
Control Test		0.079	02:17
Air Blank		0.000	02:17
Subject Sample	# 1	0.123	02:18
Breath Volume		1.687	02:18
Air Blank		0.000	02:19
Air Blank		0.000	02:21
Subject Sample	# 2	0.119	02:21
Breath Volume		2.027	02:21
Air Blank		0.000	02:22
Control Test		0.079	02:22
Air Blank		0.000	02:23
Diagnostics Check		OK	02:23

Data Download Date/Time: 02/01/2016 20:53

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,REBECCA,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PECYLAK,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 05:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:25

SUBJECT'S LAST NAME: MURATORI

DATE/TIME OF TEST: 01/22/2016 07:00:43

Diagnostics Check		OK	06:53
Air Blank		0.000	06:54
Control Test		0.079	06:54
Air Blank		0.000	06:55
Subject Sample	# 1	0.000	06:56
Breath Volume		1.207	06:56
Air Blank		0.000	06:56
Air Blank		0.000	06:58
Subject Sample	# 2	0.000	06:59
Breath Volume		1.105	06:59
Air Blank		0.000	06:59
Control Test		0.079	07:00
Air Blank		0.000	07:00
Diagnostics Check		OK	07:00

Data Download Date/Time: 02/01/2016 20:53

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DE JESUS,V
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: FLORES

DATE/TIME OF TEST: 01/23/2016 01:09:39

Diagnostics Check		OK	01:00
Air Blank		0.000	01:00
Control Test		0.079	01:01
Air Blank		0.000	01:01
Subject Sample	# 1	0.142	01:04
Breath Volume		2.242	01:04
Air Blank		0.000	01:05
Air Blank		0.000	01:06
Subject Sample	# 2	0.141	01:08
Breath Volume		1.644	01:08
Air Blank		0.000	01:08
Control Test		0.078	01:09
Air Blank		0.000	01:09
Diagnostics Check		OK	01:09

Data Download Date/Time: 02/01/2016 20:53

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: CRAGG,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:47

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 01/23/2016 04:14:14

Diagnostics Check		OK	04:07
Air Blank		0.000	04:08
Control Test		0.079	04:08
Air Blank		0.000	04:08
Subject Sample	# 1	0.082	04:09
Breath Volume		1.257	04:09
Air Blank		0.000	04:10
Air Blank		0.000	04:12
Subject Sample	# 2	0.082	04:12
Breath Volume		1.472	04:12
Air Blank		0.000	04:13
Control Test		0.078	04:13
Air Blank		0.000	04:14
Diagnostics Check		OK	04:14

Data Download Date/Time: 02/01/2016 20:53

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: CRAGG,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 05:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:03

SUBJECT'S LAST NAME: STACEY

DATE/TIME OF TEST: 01/23/2016 06:28:08

Diagnostics Check		OK	06:23	
Air Blank		0.000	06:24	
Control Test		0.079	06:24	
Air Blank		0.000	06:24	
Subject Sample	# 1		06:26	Subject Test Refused
Breath Volume		0.000	06:26	
Air Blank		0.000	06:27	
Control Test		0.079	06:27	
Air Blank		0.000	06:28	
Diagnostics Check		OK	06:28	

Data Download Date/Time: 02/01/2016 20:54

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DE JESUS,V
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 19:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:30

SUBJECT'S LAST NAME: DUMONT

DATE/TIME OF TEST: 01/23/2016 21:01:54

Diagnostics Check		OK	20:50	
Air Blank		0.000	20:50	
Control Test		0.079	20:51	
Air Blank		0.000	20:51	
Subject Sample	# 1		20:54	No Sample Provided
Breath Volume		0.000	20:54	
Air Blank		0.000	20:55	
Air Blank		0.000	20:57	
Subject Sample	# 2		21:00	No Sample Provided
Breath Volume		0.000	21:00	
Air Blank		0.000	21:00	
Control Test		0.079	21:01	
Air Blank		0.000	21:01	
Diagnostics Check		OK	21:01	

Data Download Date/Time: 02/01/2016 20:54

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY,PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ANGELICCHI,J
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: MCWHORTER

DATE/TIME OF TEST: 01/24/2016 01:28:23

Diagnostics Check		OK	01:20
Air Blank		0.000	01:21
Control Test		0.079	01:21
Air Blank		0.000	01:22
Subject Sample	# 1	0.141	01:23
Breath Volume		1.566	01:23
Air Blank		0.000	01:23
Air Blank		0.000	01:25
Subject Sample	# 2	0.133	01:26
Breath Volume		2.617	01:26
Air Blank		0.000	01:27
Control Test		0.079	01:27
Air Blank		0.000	01:28
Diagnostics Check		OK	01:28

Data Download Date/Time: 02/01/2016 20:54

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: NIEUWENDAM,R
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: TURRISI

DATE/TIME OF TEST: 01/24/2016 01:56:48

Diagnostics Check		OK	01:50
Air Blank		0.000	01:50
Control Test		0.079	01:51
Air Blank		0.000	01:51
Subject Sample	# 1	0.079	01:52
Breath Volume		1.566	01:52
Air Blank		0.000	01:52
Air Blank		0.000	01:54
Subject Sample	# 2	0.081	01:55
Breath Volume		2.199	01:55
Air Blank		0.000	01:55
Control Test		0.078	01:56
Air Blank		0.000	01:56
Diagnostics Check		OK	01:56

Data Download Date/Time: 02/01/2016 20:54

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GOFF,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:56

SUBJECT'S LAST NAME: ADAMS

DATE/TIME OF TEST: 01/24/2016 03:22:52

Diagnostics Check		OK	03:16
Air Blank		0.000	03:16
Control Test		0.079	03:16
Air Blank		0.000	03:17
Subject Sample	# 1	0.282	03:18
Breath Volume		5.140	03:18
Air Blank		0.000	03:18
Air Blank		0.000	03:20
Subject Sample	# 2	0.280	03:21
Breath Volume		5.300	03:21
Air Blank		0.000	03:21
Control Test		0.080	03:22
Air Blank		0.000	03:22
Diagnostics Check		OK	03:22

Data Download Date/Time: 02/01/2016 20:54

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:02

SUBJECT'S LAST NAME: ALVAREZ

DATE/TIME OF TEST: 01/24/2016 23:29:01

Diagnostics Check		OK	23:22
Air Blank		0.000	23:22
Control Test		0.079	23:23
Air Blank		0.000	23:23
Subject Sample	# 1	0.144	23:24
Breath Volume		2.710	23:24
Air Blank		0.000	23:24
Air Blank		0.000	23:26
Subject Sample	# 2	0.147	23:27
Breath Volume		2.542	23:27
Air Blank		0.000	23:28
Control Test		0.079	23:28
Air Blank		0.000	23:28
Diagnostics Check		OK	23:29

Data Download Date/Time: 02/01/2016 20:54

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BLANCHARD,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:43

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 01/25/2016 01:11:10

Diagnostics Check		OK	01:04
Air Blank		0.000	01:04
Control Test		0.080	01:04
Air Blank		0.000	01:05
Subject Sample	# 1	0.092	01:06
Breath Volume		1.218	01:06
Air Blank		0.000	01:07
Air Blank		0.000	01:09
Subject Sample	# 2	0.092	01:09
Breath Volume		1.265	01:09
Air Blank		0.000	01:10
Control Test		0.080	01:10
Air Blank		0.000	01:11
Diagnostics Check		OK	01:11

Data Download Date/Time: 02/01/2016 20:54

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY,PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FRANKLIN,R
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:22

SUBJECT'S LAST NAME: ACEVEDO

DATE/TIME OF TEST: 01/25/2016 01:49:59

Diagnostics Check		OK	01:43
Air Blank		0.000	01:43
Control Test		0.079	01:43
Air Blank		0.000	01:44
Subject Sample	# 1	0.165	01:45
Breath Volume		1.339	01:45
Air Blank		0.000	01:45
Air Blank		0.000	01:47
Subject Sample	# 2	0.151	01:48
Breath Volume		2.406	01:48
Air Blank		0.000	01:49
Control Test		0.079	01:49
Air Blank		0.000	01:49
Diagnostics Check		OK	01:49

Data Download Date/Time: 02/01/2016 20:54

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MALONE,C
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:23

SUBJECT'S LAST NAME: LAWRENCE

DATE/TIME OF TEST: 01/26/2016 03:50:23

Diagnostics Check		OK	03:43
Air Blank		0.000	03:44
Control Test		0.079	03:44
Air Blank		0.000	03:45
Subject Sample	# 1	0.222	03:45
Breath Volume		1.257	03:45
Air Blank		0.000	03:46
Air Blank		0.000	03:48
Subject Sample	# 2	0.211	03:48
Breath Volume		1.281	03:48
Air Blank		0.000	03:49
Control Test		0.078	03:49
Air Blank		0.000	03:50
Diagnostics Check		OK	03:50

Data Download Date/Time: 02/01/2016 20:54

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WOODEN,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:11

SUBJECT'S LAST NAME: BENNOCH

DATE/TIME OF TEST: 01/26/2016 22:37:54

Diagnostics Check		OK	22:31
Air Blank		0.000	22:31
Control Test		0.079	22:32
Air Blank		0.000	22:32
Subject Sample	# 1	0.109	22:33
Breath Volume		2.710	22:33
Air Blank		0.000	22:33
Air Blank		0.000	22:35
Subject Sample	# 2	0.108	22:36
Breath Volume		2.636	22:36
Air Blank		0.000	22:36
Control Test		0.078	22:37
Air Blank		0.000	22:37
Diagnostics Check		OK	22:37

Data Download Date/Time: 02/01/2016 20:54

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARLAUG,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:53

SUBJECT'S LAST NAME: BOTHWELL JR

DATE/TIME OF TEST: 01/28/2016 02:20:08

Diagnostics Check		OK	02:13
Air Blank		0.000	02:14
Control Test		0.079	02:14
Air Blank		0.000	02:14
Subject Sample	# 1	0.108	02:15
Breath Volume		2.710	02:15
Air Blank		0.000	02:16
Air Blank		0.000	02:18
Subject Sample	# 2	0.107	02:18
Breath Volume		3.644	02:18
Air Blank		0.000	02:19
Control Test		0.079	02:19
Air Blank		0.000	02:20
Diagnostics Check		OK	02:20

Data Download Date/Time: 02/01/2016 20:54

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SARRO,TAYLOR,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VAN-TREESE,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:43

SUBJECT'S LAST NAME: BANKS

DATE/TIME OF TEST: 01/29/2016 02:09:43

Diagnostics Check		OK	02:03
Air Blank		0.000	02:03
Control Test		0.079	02:03
Air Blank		0.000	02:04
Subject Sample	# 1	0.127	02:04
Breath Volume		2.050	02:04
Air Blank		0.000	02:05
Air Blank		0.000	02:07
Subject Sample	# 2	0.142	02:08
Breath Volume		2.074	02:08
Air Blank		0.000	02:08
Control Test		0.079	02:09
Air Blank		0.000	02:09
Diagnostics Check		OK	02:09

Data Download Date/Time: 02/01/2016 20:54

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:46

SUBJECT'S LAST NAME: HORNE

DATE/TIME OF TEST: 01/29/2016 05:13:17

Diagnostics Check		OK	05:06
Air Blank		0.000	05:07
Control Test		0.079	05:07
Air Blank		0.000	05:08
Subject Sample	# 1	0.092	05:08
Breath Volume		2.582	05:08
Air Blank		0.000	05:09
Air Blank		0.000	05:11
Subject Sample	# 2	0.090	05:11
Breath Volume		2.492	05:11
Air Blank		0.000	05:12
Control Test		0.080	05:12
Air Blank		0.000	05:13
Diagnostics Check		OK	05:13

Data Download Date/Time: 02/01/2016 20:54

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY,PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MALONE,C
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:44

SUBJECT'S LAST NAME: BREIT

DATE/TIME OF TEST: 01/30/2016 00:10:47

Diagnostics Check		OK	00:04
Air Blank		0.000	00:04
Control Test		0.079	00:05
Air Blank		0.000	00:05
Subject Sample	# 1	0.000	00:06
Breath Volume		3.593	00:06
Air Blank		0.000	00:06
Air Blank		0.000	00:08
Subject Sample	# 2	0.000	00:09
Breath Volume		3.234	00:09
Air Blank		0.000	00:09
Control Test		0.080	00:10
Air Blank		0.000	00:10
Diagnostics Check		OK	00:10

Data Download Date/Time: 02/01/2016 20:54

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DEJESUS-COLON,S
ARRESTING OFFICER AGENCY: Temple Terrace PD
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:58

SUBJECT'S LAST NAME: RAMKISSOON

DATE/TIME OF TEST: 01/30/2016 04:25:57

Diagnostics Check		OK	04:18
Air Blank		0.000	04:19
Control Test		0.079	04:19
Air Blank		0.000	04:20
Subject Sample	# 1	0.183	04:21
Breath Volume		1.718	04:21
Air Blank		0.000	04:21
Air Blank		0.000	04:23
Subject Sample	# 2	0.186	04:24
Breath Volume		1.726	04:24
Air Blank		0.000	04:25
Control Test		0.079	04:25
Air Blank		0.000	04:25
Diagnostics Check		OK	04:25

Data Download Date/Time: 02/01/2016 20:54

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNYDER,MELONIE,E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: CORTES,P
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 17:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:25

SUBJECT'S LAST NAME: HORTON

DATE/TIME OF TEST: 01/30/2016 17:51:20

Diagnostics Check		OK	17:45
Air Blank		0.000	17:45
Control Test		0.079	17:46
Air Blank		0.000	17:46
Subject Sample	# 1	0.021	17:46
Breath Volume		1.253	17:46
Air Blank		0.000	17:47
Air Blank		0.000	17:49
Subject Sample	# 2	0.020	17:49
Breath Volume		1.710	17:49
Air Blank		0.000	17:50
Control Test		0.079	17:50
Air Blank		0.000	17:51
Diagnostics Check		OK	17:51

Data Download Date/Time: 02/01/2016 20:54

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:58

SUBJECT'S LAST NAME: STEVENSON

DATE/TIME OF TEST: 01/30/2016 21:24:41

Diagnostics Check		OK	21:18
Air Blank		0.000	21:18
Control Test		0.079	21:19
Air Blank		0.000	21:19
Subject Sample	# 1	0.208	21:20
Breath Volume		1.457	21:20
Air Blank		0.000	21:20
Air Blank		0.000	21:22
Subject Sample	# 2	0.194	21:23
Breath Volume		1.414	21:23
Air Blank		0.000	21:23
Control Test		0.078	21:24
Air Blank		0.000	21:24
Diagnostics Check		OK	21:24

Data Download Date/Time: 02/01/2016 20:54

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: CRAGG,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:34

SUBJECT'S LAST NAME: GUTTMAN

DATE/TIME OF TEST: 01/31/2016 02:23:55

Diagnostics Check		OK	02:21	
Air Blank		0.000	02:21	
Control Test		0.078	02:21	
Air Blank		0.000	02:22	
Subject Sample	# 1		02:22	Subject Test Refused
Breath Volume		0.000	02:22	
Air Blank		0.000	02:23	
Control Test		0.079	02:23	
Air Blank		0.000	02:23	
Diagnostics Check		OK	02:23	

Data Download Date/Time: 02/01/2016 20:54

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: AGUAYO,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: RYDER

DATE/TIME OF TEST: 01/31/2016 03:34:50

Diagnostics Check		OK	03:32	
Air Blank		0.000	03:32	
Control Test		0.079	03:32	
Air Blank		0.000	03:33	
Subject Sample	# 1		03:33	Subject Test Refused
Breath Volume		0.000	03:33	
Air Blank		0.000	03:33	
Control Test		0.079	03:34	
Air Blank		0.000	03:34	
Diagnostics Check		OK	03:34	

Data Download Date/Time: 02/01/2016 20:54

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005248

INSTRUMENT REGISTERED TO: FWC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIAZ,ALEXANDER,
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: MERT,SERDAR
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 12/20/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1451362
DRY GAS EXPIRY DATE: 01/31/2016
VIOLATION CODE: BUI
ARREST TIME: 18:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:29

SUBJECT'S LAST NAME: SALAZAR

DATE/TIME OF TEST: 01/10/2016 19:24:35

Diagnostics Check		OK	19:15
Air Blank		0.000	19:15
Control Test		0.081	19:15
Air Blank		0.000	19:16
Subject Sample	# 1	0.139	19:17
Breath Volume		2.585	19:17
Air Blank		0.000	19:17
Air Blank		0.000	19:19
Subject Sample	# 2	0.118	19:19
Breath Volume		1.867	19:19
Air Blank		0.000	19:20
Air Blank		0.000	19:22
Subject Sample	# 3	0.129	19:22
Breath Volume		1.621	19:22
Air Blank		0.000	19:23
Control Test		0.081	19:23
Air Blank		0.000	19:24
Diagnostics Check		OK	19:24

Data Download Date/Time: 01/26/2016 18:20

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRITZ,STEPHEN,A
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: PRITZ,STEPHEN
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 12/21/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 16014080A2
DRY GAS EXPIRY DATE: 07/05/2016
VIOLATION CODE: DUI
ARREST TIME: 02:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:49

SUBJECT'S LAST NAME: PEDONE

DATE/TIME OF TEST: 01/01/2016 03:15:13

Diagnostics Check		OK	03:09
Air Blank		0.000	03:09
Control Test		0.081	03:09
Air Blank		0.000	03:10
Subject Sample	# 1	0.117	03:10
Breath Volume		1.625	03:10
Air Blank		0.000	03:11
Air Blank		0.000	03:13
Subject Sample	# 2	0.108	03:13
Breath Volume		1.226	03:13
Air Blank		0.000	03:14
Control Test		0.081	03:14
Air Blank		0.000	03:15
Diagnostics Check		OK	03:15

Data Download Date/Time: 01/22/2016 09:49

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAOLELLA,MICHAEL,P
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: WITHERELL,B
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 12/21/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 16014080A2
DRY GAS EXPIRY DATE: 07/05/2016
VIOLATION CODE: DUI
ARREST TIME: 04:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:15

SUBJECT'S LAST NAME: PERKINS

DATE/TIME OF TEST: 01/01/2016 05:39:29

Diagnostics Check		OK	05:35	
Air Blank		0.000	05:36	
Control Test		0.082	05:36	
Air Blank		0.000	05:36	
Subject Sample	# 1		05:38	Subject Test Refused
Breath Volume		0.000	05:38	
Air Blank		0.000	05:38	
Control Test		0.081	05:38	
Air Blank		0.000	05:39	
Diagnostics Check		OK	05:39	

Data Download Date/Time: 01/22/2016 09:49

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRITZ,STEPHEN,A
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: PRITZ,STEPHEN
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 12/21/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 16014080A2
DRY GAS EXPIRY DATE: 07/05/2016
VIOLATION CODE: DUI
ARREST TIME: 00:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:48

SUBJECT'S LAST NAME: HUGHES

DATE/TIME OF TEST: 01/02/2016 01:15:15

Diagnostics Check		OK	01:09
Air Blank		0.000	01:09
Control Test		0.081	01:09
Air Blank		0.000	01:10
Subject Sample	# 1	0.118	01:10
Breath Volume		2.582	01:10
Air Blank		0.000	01:11
Air Blank		0.000	01:13
Subject Sample	# 2	0.116	01:13
Breath Volume		2.832	01:13
Air Blank		0.000	01:14
Control Test		0.080	01:14
Air Blank		0.000	01:15
Diagnostics Check		OK	01:15

Data Download Date/Time: 01/22/2016 09:49

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEIRCE,HAROLD,F
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: SCHAUB,ERIC
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/21/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 16014080A2
DRY GAS EXPIRY DATE: 07/05/2016
VIOLATION CODE: DUI
ARREST TIME: 08:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:10

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 01/03/2016 09:37:47

Diagnostics Check		OK	09:30
Air Blank		0.000	09:31
Control Test		0.082	09:31
Air Blank		0.000	09:32
Subject Sample	# 1	0.174	09:32
Breath Volume		3.750	09:32
Air Blank		0.000	09:33
Air Blank		0.000	09:35
Subject Sample	# 2	0.169	09:36
Breath Volume		3.195	09:36
Air Blank		0.000	09:36
Control Test		0.080	09:37
Air Blank		0.000	09:37
Diagnostics Check		OK	09:37

Data Download Date/Time: 01/22/2016 09:49

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BEETZ,JUSTIN,W
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: WEEKS,JOHNATHAN
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 12/21/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 16014080A2
DRY GAS EXPIRY DATE: 07/05/2016
VIOLATION CODE: DUI
ARREST TIME: 03:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:12

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 01/04/2016 05:35:47

Diagnostics Check		OK	05:32	
Air Blank		0.000	05:33	
Control Test		0.083	05:33	
Air Blank		0.000	05:34	
Subject Sample	# 1		05:34	Subject Test Refused
Breath Volume		0.000	05:34	
Air Blank		0.000	05:34	
Control Test		0.082	05:35	
Air Blank		0.000	05:35	
Diagnostics Check		OK	05:35	

Data Download Date/Time: 01/22/2016 09:49

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BEETZ,JUSTIN,W
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: REID,WAREN
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 12/21/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 16014080A2
DRY GAS EXPIRY DATE: 07/05/2016
VIOLATION CODE: DUI
ARREST TIME: 01:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:49

SUBJECT'S LAST NAME: MCCLURE

DATE/TIME OF TEST: 01/07/2016 03:13:01

Diagnostics Check		OK	03:09	
Air Blank		0.000	03:10	
Control Test		0.081	03:10	
Air Blank		0.000	03:11	
Subject Sample	# 1		03:11	Subject Test Refused
Breath Volume		0.000	03:11	
Air Blank		0.000	03:12	
Control Test		0.081	03:12	
Air Blank		0.000	03:12	
Diagnostics Check		OK	03:13	

Data Download Date/Time: 01/22/2016 09:49

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPENCER,COLLEEN,E
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: ORIGON,LANCE
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 12/21/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 16014080A2
DRY GAS EXPIRY DATE: 07/05/2016
VIOLATION CODE: DUI
ARREST TIME: 13:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:46

SUBJECT'S LAST NAME: SCOTT

DATE/TIME OF TEST: 01/08/2016 15:08:58

Diagnostics Check		OK	15:06	
Air Blank		0.000	15:06	
Control Test		0.082	15:06	
Air Blank		0.000	15:07	
Subject Sample	# 1		15:07	Subject Test Refused
Breath Volume		0.000	15:07	
Air Blank		0.000	15:08	
Control Test		0.081	15:08	
Air Blank		0.000	15:08	
Diagnostics Check		OK	15:08	

Data Download Date/Time: 01/22/2016 09:49

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANTORO,ROBERT,
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: SANTORO,ROBERT
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 12/21/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 16014080A2
DRY GAS EXPIRY DATE: 07/05/2016
VIOLATION CODE: DUI
ARREST TIME: 18:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:40

SUBJECT'S LAST NAME: BLAIR

DATE/TIME OF TEST: 01/13/2016 19:03:25

Diagnostics Check		OK	19:00	
Air Blank		0.000	19:01	
Control Test		0.083	19:01	
Air Blank		0.000	19:01	
Subject Sample	# 1		19:02	Subject Test Refused
Breath Volume		0.000	19:02	
Air Blank		0.000	19:02	
Control Test		0.082	19:02	
Air Blank		0.000	19:03	
Diagnostics Check		OK	19:03	

Data Download Date/Time: 01/22/2016 09:49

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANTORO,ROBERT,T
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: SANTORO,ROBERT
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 12/21/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 16014080A2
DRY GAS EXPIRY DATE: 07/05/2016
VIOLATION CODE: DUI
ARREST TIME: 19:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:25

SUBJECT'S LAST NAME: FRANCESCO

DATE/TIME OF TEST: 01/14/2016 20:52:53

Diagnostics Check		OK	20:45
Air Blank		0.000	20:46
Control Test		0.081	20:46
Air Blank		0.000	20:46
Subject Sample	# 1	0.129	20:47
Breath Volume		1.257	20:47
Air Blank		0.000	20:48
Air Blank		0.000	20:50
Subject Sample	# 2	0.129	20:51
Breath Volume		2.042	20:51
Air Blank		0.000	20:51
Control Test		0.080	20:52
Air Blank		0.000	20:52
Diagnostics Check		OK	20:52

Data Download Date/Time: 01/22/2016 09:49

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORROW,TRICIA,J
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: FREDRICKSEN,S
ARRESTING OFFICER AGENCY: Brooksville PD
LAST AGENCY INSPECTION: 12/21/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 16014080A2
DRY GAS EXPIRY DATE: 07/05/2016
VIOLATION CODE: DUI
ARREST TIME: 00:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:52

SUBJECT'S LAST NAME: BARICH

DATE/TIME OF TEST: 01/15/2016 01:19:38

Diagnostics Check		OK	01:12
Air Blank		0.000	01:13
Control Test		0.082	01:13
Air Blank		0.000	01:14
Subject Sample	# 1	0.168	01:14
Breath Volume		1.183	01:14
Air Blank		0.000	01:15
Air Blank		0.000	01:17
Subject Sample	# 2	0.164	01:18
Breath Volume		2.609	01:18
Air Blank		0.000	01:18
Control Test		0.082	01:19
Air Blank		0.000	01:19
Diagnostics Check		OK	01:19

Data Download Date/Time: 01/22/2016 09:49

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRITZ,STEPHEN,A
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: PRITZ,STEPHEN
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 12/21/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 16014080A2
DRY GAS EXPIRY DATE: 07/05/2016
VIOLATION CODE: DUI
ARREST TIME: 16:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:12

SUBJECT'S LAST NAME: INGEMI

DATE/TIME OF TEST: 01/19/2016 17:38:47

Diagnostics Check		OK	17:32
Air Blank		0.000	17:33
Control Test		0.084	17:33
Air Blank		0.000	17:33
Subject Sample	# 1	0.000	17:34
Breath Volume		2.476	17:34
Air Blank		0.000	17:34
Air Blank		0.000	17:36
Subject Sample	# 2	0.000	17:37
Breath Volume		3.007	17:37
Air Blank		0.000	17:37
Control Test		0.083	17:38
Air Blank		0.000	17:38
Diagnostics Check		OK	17:38

Data Download Date/Time: 01/22/2016 09:49

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORROW,T,J
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: HOAGLAN,B
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 16014080A2
DRY GAS EXPIRY DATE: 07/05/2016
VIOLATION CODE: DUI
ARREST TIME: 18:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:54

SUBJECT'S LAST NAME: WORTHINGTON

DATE/TIME OF TEST: 01/23/2016 19:21:27

Diagnostics Check		OK	19:14
Air Blank		0.000	19:15
Control Test		0.083	19:15
Air Blank		0.000	19:15
Subject Sample	# 1	0.178	19:16
Breath Volume		1.824	19:16
Air Blank		0.000	19:17
Air Blank		0.000	19:19
Subject Sample	# 2	0.178	19:19
Breath Volume		2.156	19:19
Air Blank		0.000	19:20
Control Test		0.081	19:20
Air Blank		0.000	19:21
Diagnostics Check		OK	19:21

Data Download Date/Time: 02/20/2016 09:09

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANTORO,ROBERT,T
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: SANTORO,ROBERT
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 01/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 16014080A2
DRY GAS EXPIRY DATE: 07/05/2016
VIOLATION CODE: DUI
ARREST TIME: 20:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: LAVALLE

DATE/TIME OF TEST: 01/23/2016 21:17:56

Diagnostics Check		OK	21:10
Air Blank		0.000	21:10
Control Test		0.082	21:11
Air Blank		0.000	21:11
Subject Sample	# 1	0.128	21:13
Breath Volume		1.203	21:13
Air Blank		0.000	21:14
Air Blank		0.000	21:15
Subject Sample	# 2	0.136	21:16
Breath Volume		1.308	21:16
Air Blank		0.000	21:17
Control Test		0.081	21:17
Air Blank		0.000	21:17
Diagnostics Check		OK	21:17

Data Download Date/Time: 02/20/2016 09:09

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BEETZ,JUSTIN,W
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: REID,WARREN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 16014080A2
DRY GAS EXPIRY DATE: 07/05/2016
VIOLATION CODE: DUI
ARREST TIME: 20:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:35

SUBJECT'S LAST NAME: GOODWIN

DATE/TIME OF TEST: 01/26/2016 22:09:33

Diagnostics Check		OK	21:55	
Air Blank		0.000	21:55	
Control Test		0.082	21:55	
Air Blank		0.000	21:56	
Subject Sample	# 1	0.000	21:59	Volume Not Met
Breath Volume		0.992	21:59	
Air Blank		0.000	22:00	
Air Blank		0.000	22:02	
Subject Sample	# 2	0.000	22:03	
Breath Volume		1.210	22:03	
Air Blank		0.000	22:03	
Air Blank		0.000	22:05	
Subject Sample	# 3	0.000	22:08	
Breath Volume		1.332	22:08	
Air Blank		0.000	22:08	
Control Test		0.081	22:08	
Air Blank		0.000	22:09	
Diagnostics Check		OK	22:09	

Data Download Date/Time: 02/20/2016 09:09

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANTORO,ROBERT,T
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: SANTORO,ROBERT
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 01/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 16014080A2
DRY GAS EXPIRY DATE: 07/05/2016
VIOLATION CODE: DUI
ARREST TIME: 22:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:22

SUBJECT'S LAST NAME: READO

DATE/TIME OF TEST: 01/29/2016 23:27:44

Diagnostics Check		OK	23:24	
Air Blank		0.000	23:25	
Control Test		0.082	23:25	
Air Blank		0.000	23:26	
Subject Sample	# 1		23:26	Subject Test Refused
Breath Volume		0.000	23:26	
Air Blank		0.000	23:26	
Control Test		0.082	23:27	
Air Blank		0.000	23:27	
Diagnostics Check		OK	23:27	

Data Download Date/Time: 02/20/2016 09:09

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANTORO,ROBERT,T
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: SANTORO,ROBERT
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 01/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 16014080A2
DRY GAS EXPIRY DATE: 07/05/2016
VIOLATION CODE: DUI
ARREST TIME: 22:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:27

SUBJECT'S LAST NAME: READO

DATE/TIME OF TEST: 01/29/2016 23:53:33

Diagnostics Check		OK	23:47
Air Blank		0.000	23:47
Control Test		0.081	23:48
Air Blank		0.000	23:48
Subject Sample	# 1	0.153	23:49
Breath Volume		1.636	23:49
Air Blank		0.000	23:49
Air Blank		0.000	23:51
Subject Sample	# 2	0.139	23:52
Breath Volume		1.417	23:52
Air Blank		0.000	23:52
Control Test		0.081	23:52
Air Blank		0.000	23:53
Diagnostics Check		OK	23:53

Data Download Date/Time: 02/20/2016 09:09

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005337

INSTRUMENT REGISTERED TO: PERRY POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CANNON,CHAD,M
BREATH TEST OPERATOR AGENCY: Perry PD
ARRESTING OFFICER: HAYDEN,BRYAN
ARRESTING OFFICER AGENCY: Taylor County SO
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: AG425202
DRY GAS EXPIRY DATE: 09/09/2016
VIOLATION CODE: DUI
ARREST TIME: 19:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:19

SUBJECT'S LAST NAME: COLLINS

DATE/TIME OF TEST: 01/15/2016 19:48:25

Diagnostics Check		OK	19:42
Air Blank		0.000	19:42
Control Test		0.080	19:42
Air Blank		0.000	19:43
Subject Sample	# 1	0.318	19:43
Breath Volume		1.199	19:43
Air Blank		0.000	19:44
Air Blank		0.000	19:46
Subject Sample	# 2	0.308	19:46
Breath Volume		2.250	19:46
Air Blank		0.000	19:47
Control Test		0.079	19:47
Air Blank		0.000	19:48
Diagnostics Check		OK	19:48

Data Download Date/Time: 01/29/2016 10:24

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005337

INSTRUMENT REGISTERED TO: PERRY POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICKETSON,JAY,R
BREATH TEST OPERATOR AGENCY: Perry PD
ARRESTING OFFICER: RICKETSON,J
ARRESTING OFFICER AGENCY: Perry PD
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG425202
DRY GAS EXPIRY DATE: 09/09/2016
VIOLATION CODE: DUI
ARREST TIME: 11:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:00

SUBJECT'S LAST NAME: FREEMAN

DATE/TIME OF TEST: 01/25/2016 11:55:35

Diagnostics Check	OK	11:54	
Air Blank		11:55	RFI Detect
Air Blank	0.000	11:55	

Data Download Date/Time: 01/29/2016 10:24

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005337

INSTRUMENT REGISTERED TO: PERRY POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICKETSON,J,R
BREATH TEST OPERATOR AGENCY: Perry PD
ARRESTING OFFICER: RICKETSON,J
ARRESTING OFFICER AGENCY: Perry PD
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG425202
DRY GAS EXPIRY DATE: 09/09/2016
VIOLATION CODE: DUI
ARREST TIME: 11:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:00

SUBJECT'S LAST NAME: FREEMAN

DATE/TIME OF TEST: 01/25/2016 12:03:37

Diagnostics Check		OK	11:57
Air Blank		0.000	11:58
Control Test		0.081	11:58
Air Blank		0.000	11:58
Subject Sample	# 1	0.000	11:59
Breath Volume		4.382	11:59
Air Blank		0.000	11:59
Air Blank		0.000	12:01
Subject Sample	# 2	0.000	12:02
Breath Volume		3.878	12:02
Air Blank		0.000	12:02
Control Test		0.080	12:03
Air Blank		0.000	12:03
Diagnostics Check		OK	12:03

Data Download Date/Time: 01/29/2016 10:24

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005337

INSTRUMENT REGISTERED TO: PERRY POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICKETSON,J,R
BREATH TEST OPERATOR AGENCY: Perry PD
ARRESTING OFFICER: RICKETSON,J
ARRESTING OFFICER AGENCY: Perry PD
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG425202
DRY GAS EXPIRY DATE: 09/09/2016
VIOLATION CODE: ADMIN
ARREST TIME: 11:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:50

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 01/26/2016 12:31:22

Diagnostics Check		OK	12:25
Air Blank		0.000	12:25
Control Test		0.081	12:25
Air Blank		0.000	12:26
Subject Sample	# 1	0.019	12:27
Breath Volume		2.238	12:27
Air Blank		0.000	12:27
Air Blank		0.000	12:29
Subject Sample	# 2	0.018	12:29
Breath Volume		3.792	12:29
Air Blank		0.000	12:30
Control Test		0.080	12:30
Air Blank		0.000	12:31
Diagnostics Check		OK	12:31

Data Download Date/Time: 01/29/2016 10:24

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005337

INSTRUMENT REGISTERED TO: PERRY POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MURPHY,TIMOTHY,L
BREATH TEST OPERATOR AGENCY: Perry PD
ARRESTING OFFICER: MURPHY,TIMOTHY
ARRESTING OFFICER AGENCY: Perry PD
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG425202
DRY GAS EXPIRY DATE: 09/09/2016
VIOLATION CODE: ADMIN
ARREST TIME: 09:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:10

SUBJECT'S LAST NAME: LANGFORD

DATE/TIME OF TEST: 01/28/2016 09:37:15

Diagnostics Check		OK	09:31
Air Blank		0.000	09:31
Control Test		0.081	09:31
Air Blank		0.000	09:32
Subject Sample	# 1	0.000	09:32
Breath Volume		2.054	09:32
Air Blank		0.000	09:33
Air Blank		0.000	09:35
Subject Sample	# 2	0.000	09:35
Breath Volume		2.488	09:35
Air Blank		0.000	09:36
Control Test		0.078	09:36
Air Blank		0.000	09:37
Diagnostics Check		OK	09:37

Data Download Date/Time: 01/29/2016 10:24

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLARK,BRENT,D
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CLARK,BRENT
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/25/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31113080A1
DRY GAS EXPIRY DATE: 12/01/2015
VIOLATION CODE: DUI
ARREST TIME: 20:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:09

SUBJECT'S LAST NAME: CHESHIRE

DATE/TIME OF TEST: 01/24/2016 21:46:28

Diagnostics Check		OK	21:38	
Air Blank		0.000	21:38	
Control Test		0.078	21:38	
Air Blank		0.000	21:39	
Subject Sample	# 1	0.186	21:42	Volume Not Met
Breath Volume		0.941	21:42	
Air Blank		0.000	21:43	
Air Blank		0.000	21:44	
Subject Sample	# 2		21:45	Subject Test Refused
Breath Volume		0.000	21:45	
Air Blank		0.000	21:45	
Control Test		0.079	21:45	
Air Blank		0.000	21:46	
Diagnostics Check		OK	21:46	

Data Download Date/Time: 01/30/2016 15:02

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOULD,LEONARD,
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: GOULD,LEONARD
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 01/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29614080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 22:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: HERRERA

DATE/TIME OF TEST: 01/28/2016 23:01:43

Diagnostics Check	OK	23:00	
Air Blank	0.000	23:00	
Control Test	0.082	23:01	Interferent Detect
Air Blank	0.000	23:01	

Data Download Date/Time: 02/17/2016 12:00

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOULD,LEONARD,
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: GOULD,LEONARD
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 01/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29614080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 22:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: HERRERA

DATE/TIME OF TEST: 01/28/2016 23:06:11

Diagnostics Check		OK	23:03	
Air Blank		0.000	23:03	
Control Test		0.080	23:04	
Air Blank		0.000	23:04	
Subject Sample	# 1	0.168	23:04	
Breath Volume		1.730	23:04	
Air Blank			23:05	Purge Fail
Air Blank		0.000	23:06	

Data Download Date/Time: 02/17/2016 12:00

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005537

INSTRUMENT REGISTERED TO: DAYTONA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FORREST,RYAN,A
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: LIMA,MICHAEL
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 12/18/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1669440
DRY GAS EXPIRY DATE: 04/19/2017
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: ALLEN

DATE/TIME OF TEST: 01/11/2016 03:29:15

Diagnostics Check		OK	03:21
Air Blank		0.000	03:21
Control Test		0.080	03:22
Air Blank		0.000	03:22
Subject Sample	# 1	0.131	03:24
Breath Volume		2.304	03:24
Air Blank		0.000	03:25
Air Blank		0.000	03:26
Subject Sample	# 2	0.132	03:27
Breath Volume		1.410	03:27
Air Blank		0.000	03:28
Control Test		0.078	03:28
Air Blank		0.000	03:29
Diagnostics Check		OK	03:29

Data Download Date/Time: 01/27/2016 22:32

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005537

INSTRUMENT REGISTERED TO: DAYTONA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUNNE,JUSTIN,R
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: WARD,JAMES
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 12/18/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1669440
DRY GAS EXPIRY DATE: 04/19/2017
VIOLATION CODE: DUI
ARREST TIME: 03:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:38

SUBJECT'S LAST NAME: ANDUZE

DATE/TIME OF TEST: 01/15/2016 04:04:02

Diagnostics Check		OK	04:01	
Air Blank		0.000	04:01	
Control Test		0.080	04:01	
Air Blank		0.000	04:02	
Subject Sample	# 1		04:02	Subject Test Refused
Breath Volume		0.000	04:02	
Air Blank		0.000	04:03	
Control Test		0.080	04:03	
Air Blank		0.000	04:03	
Diagnostics Check		OK	04:04	

Data Download Date/Time: 01/27/2016 22:32

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005537

INSTRUMENT REGISTERED TO: DAYTONA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAMPE,CORTLAND,B
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: MILLS,RESHARD
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 12/18/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1669440
DRY GAS EXPIRY DATE: 04/19/2017
VIOLATION CODE: DUI
ARREST TIME: 19:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:20

SUBJECT'S LAST NAME: TAMER

DATE/TIME OF TEST: 01/16/2016 20:47:15

Diagnostics Check		OK	20:40
Air Blank		0.000	20:40
Control Test		0.080	20:41
Air Blank		0.000	20:41
Subject Sample	# 1	0.281	20:42
Breath Volume		2.308	20:42
Air Blank		0.000	20:43
Air Blank		0.000	20:44
Subject Sample	# 2	0.291	20:45
Breath Volume		2.031	20:45
Air Blank		0.000	20:46
Control Test		0.080	20:46
Air Blank		0.000	20:47
Diagnostics Check		OK	20:47

Data Download Date/Time: 01/27/2016 22:32

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SORRELL,BRIAN,J
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: SORRELL,BRIAN
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 01/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 17:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:50

SUBJECT'S LAST NAME: FALLS

DATE/TIME OF TEST: 01/19/2016 18:23:48

Diagnostics Check		OK	18:17
Air Blank		0.000	18:18
Control Test		0.080	18:18
Air Blank		0.000	18:18
Subject Sample	# 1	0.329	18:19
Breath Volume		1.210	18:19
Air Blank		0.000	18:19
Air Blank		0.000	18:21
Subject Sample	# 2	0.324	18:22
Breath Volume		1.562	18:22
Air Blank		0.000	18:22
Control Test		0.079	18:23
Air Blank		0.000	18:23
Diagnostics Check		OK	18:23

Data Download Date/Time: 02/12/2016 09:28

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VARGAS,D,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: VARGAS,D
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 01/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 01:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:49

SUBJECT'S LAST NAME: CECCHI

DATE/TIME OF TEST: 01/30/2016 02:17:45

Diagnostics Check		OK	02:12	
Air Blank		0.000	02:13	
Control Test		0.081	02:13	
Air Blank		0.000	02:14	
Subject Sample	# 1		02:16	Subject Test Refused
Breath Volume		0.000	02:16	
Air Blank		0.000	02:16	
Control Test		0.079	02:17	
Air Blank		0.000	02:17	
Diagnostics Check		OK	02:17	

Data Download Date/Time: 02/12/2016 09:28

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: CLOSIUS,RICHARD
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 12/10/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 05:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:39

SUBJECT'S LAST NAME: HUECK

DATE/TIME OF TEST: 01/01/2016 06:05:44

Diagnostics Check		OK	05:59
Air Blank		0.000	06:00
Control Test		0.078	06:00
Air Blank		0.000	06:00
Subject Sample	# 1	0.000	06:01
Breath Volume		1.257	06:01
Air Blank		0.000	06:01
Air Blank		0.000	06:03
Subject Sample	# 2	0.000	06:04
Breath Volume		2.027	06:04
Air Blank		0.000	06:04
Control Test		0.077	06:05
Air Blank		0.000	06:05
Diagnostics Check		OK	06:05

Data Download Date/Time: 01/06/2016 10:33

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 12/10/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 05:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:47

SUBJECT'S LAST NAME: DAMODARAN

DATE/TIME OF TEST: 01/01/2016 06:19:45

Diagnostics Check		OK	06:13
Air Blank		0.000	06:13
Control Test		0.078	06:14
Air Blank		0.000	06:14
Subject Sample	# 1	0.097	06:15
Breath Volume		1.804	06:15
Air Blank		0.000	06:15
Air Blank		0.000	06:17
Subject Sample	# 2	0.100	06:18
Breath Volume		2.617	06:18
Air Blank		0.000	06:18
Control Test		0.077	06:19
Air Blank		0.000	06:19
Diagnostics Check		OK	06:19

Data Download Date/Time: 01/06/2016 10:33

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VARGAS,DAEL,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 12/10/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 23:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:42

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 01/02/2016 00:15:35

Diagnostics Check		OK	00:07
Air Blank		0.000	00:08
Control Test		0.077	00:08
Air Blank		0.000	00:09
Subject Sample	# 1	0.066	00:09
Breath Volume		1.230	00:09
Air Blank		0.000	00:10
Air Blank		0.000	00:12
Subject Sample	# 2	0.070	00:13
Breath Volume		2.074	00:13
Air Blank		0.000	00:14
Control Test		0.077	00:14
Air Blank		0.000	00:15
Diagnostics Check		OK	00:15

Data Download Date/Time: 01/06/2016 10:33

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: CLOSIUS,RICHARD
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 12/10/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 04:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:40

SUBJECT'S LAST NAME: VALDES

DATE/TIME OF TEST: 01/02/2016 05:06:27

Diagnostics Check		OK	05:00
Air Blank		0.000	05:00
Control Test		0.077	05:01
Air Blank		0.000	05:01
Subject Sample	# 1	0.176	05:01
Breath Volume		2.785	05:01
Air Blank		0.000	05:02
Air Blank		0.000	05:04
Subject Sample	# 2	0.179	05:04
Breath Volume		2.593	05:04
Air Blank		0.000	05:05
Control Test		0.077	05:05
Air Blank		0.000	05:06
Diagnostics Check		OK	05:06

Data Download Date/Time: 01/06/2016 10:33

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 12/10/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 04:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:25

SUBJECT'S LAST NAME: VAUGHN

DATE/TIME OF TEST: 01/02/2016 05:56:05

Diagnostics Check		OK	05:49
Air Blank		0.000	05:49
Control Test		0.078	05:49
Air Blank		0.000	05:50
Subject Sample	# 1	0.053	05:50
Breath Volume		1.687	05:50
Air Blank		0.000	05:51
Air Blank		0.000	05:53
Subject Sample	# 2	0.050	05:54
Breath Volume		1.750	05:54
Air Blank		0.000	05:55
Control Test		0.078	05:55
Air Blank		0.000	05:56
Diagnostics Check		OK	05:56

Data Download Date/Time: 01/06/2016 10:33

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZRAK,JOSEPH,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: CRUZ,E
ARRESTING OFFICER AGENCY: Village of Pinecrest
LAST AGENCY INSPECTION: 12/10/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 00:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:02

SUBJECT'S LAST NAME: LIMA

DATE/TIME OF TEST: 01/01/2016 01:51:47

Diagnostics Check		OK	01:48	
Air Blank		0.000	01:49	
Control Test		0.078	01:49	
Air Blank		0.000	01:50	
Subject Sample	# 1		01:50	Subject Test Refused
Breath Volume		0.000	01:50	
Air Blank		0.000	01:50	
Control Test		0.078	01:51	
Air Blank		0.000	01:51	
Diagnostics Check		OK	01:51	

Data Download Date/Time: 01/06/2016 10:30

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZRAK,JOSEPH,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: AZRAK,JOSEPH
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 12/10/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 04:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:49

SUBJECT'S LAST NAME: ACEBO

DATE/TIME OF TEST: 01/01/2016 05:16:24

Diagnostics Check		OK	05:10
Air Blank		0.000	05:10
Control Test		0.078	05:11
Air Blank		0.000	05:11
Subject Sample	# 1	0.157	05:12
Breath Volume		1.230	05:12
Air Blank		0.000	05:12
Air Blank		0.000	05:14
Subject Sample	# 2	0.162	05:14
Breath Volume		2.523	05:14
Air Blank		0.000	05:15
Control Test		0.078	05:15
Air Blank		0.000	05:16
Diagnostics Check		OK	05:16

Data Download Date/Time: 01/06/2016 10:30

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZRAK,JOSEPH,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: AZRAK,JOSEPH
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 12/10/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 04:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:09

SUBJECT'S LAST NAME: ALBOR

DATE/TIME OF TEST: 01/02/2016 04:36:39

Diagnostics Check		OK	04:29
Air Blank		0.000	04:30
Control Test		0.078	04:30
Air Blank		0.000	04:31
Subject Sample	# 1	0.152	04:31
Breath Volume		3.203	04:31
Air Blank		0.000	04:32
Air Blank		0.000	04:34
Subject Sample	# 2	0.156	04:35
Breath Volume		2.878	04:35
Air Blank		0.000	04:35
Control Test		0.078	04:36
Air Blank		0.000	04:36
Diagnostics Check		OK	04:36

Data Download Date/Time: 01/06/2016 10:30

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZRAK,JOSEPH,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: CRUZ,E
ARRESTING OFFICER AGENCY: Village of Pinecrest
LAST AGENCY INSPECTION: 12/10/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 05:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:20

SUBJECT'S LAST NAME: PINEDA-RIVERA

DATE/TIME OF TEST: 01/02/2016 05:49:15

Diagnostics Check		OK	05:43
Air Blank		0.000	05:43
Control Test		0.078	05:44
Air Blank		0.000	05:44
Subject Sample	# 1	0.139	05:44
Breath Volume		2.261	05:44
Air Blank		0.000	05:45
Air Blank		0.000	05:47
Subject Sample	# 2	0.147	05:47
Breath Volume		2.312	05:47
Air Blank		0.000	05:48
Control Test		0.078	05:48
Air Blank		0.000	05:49
Diagnostics Check		OK	05:49

Data Download Date/Time: 01/06/2016 10:30

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZRAK,JOSEPH,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: AZRAK,JOSEPH
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 01/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: DESSOURCES

DATE/TIME OF TEST: 01/07/2016 03:17:07

Diagnostics Check		OK	03:14	
Air Blank		0.000	03:14	
Control Test		0.079	03:14	
Air Blank		0.000	03:15	
Subject Sample	# 1		03:15	Subject Test Refused
Breath Volume		0.000	03:15	
Air Blank		0.000	03:16	
Control Test		0.079	03:16	
Air Blank		0.000	03:17	
Diagnostics Check		OK	03:17	

Data Download Date/Time: 02/10/2016 08:55

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZRAK,JOSEPH,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: AZRAK,JOSEPH
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 01/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 03:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:46

SUBJECT'S LAST NAME: FERRER

DATE/TIME OF TEST: 01/13/2016 04:28:29

Diagnostics Check		OK	04:25	
Air Blank		0.000	04:26	
Control Test		0.080	04:26	
Air Blank		0.000	04:26	
Subject Sample	# 1		04:27	Subject Test Refused
Breath Volume		0.000	04:27	
Air Blank		0.000	04:27	
Control Test		0.079	04:27	
Air Blank		0.000	04:28	
Diagnostics Check		OK	04:28	

Data Download Date/Time: 02/10/2016 08:55

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZRAK,JOSEPH,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: AZRAK,JOSEPH
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 01/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 02:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: VELASCO-OYOLA

DATE/TIME OF TEST: 01/14/2016 02:42:15

Diagnostics Check		OK	02:35
Air Blank		0.000	02:36
Control Test		0.079	02:36
Air Blank		0.000	02:37
Subject Sample	# 1	0.185	02:37
Breath Volume		1.761	02:37
Air Blank		0.000	02:38
Air Blank		0.000	02:40
Subject Sample	# 2	0.183	02:40
Breath Volume		2.613	02:40
Air Blank		0.000	02:41
Control Test		0.079	02:41
Air Blank		0.000	02:42
Diagnostics Check		OK	02:42

Data Download Date/Time: 02/10/2016 08:55

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZRAK,JOSEPH,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: AZRAK,JOSEPH
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 01/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 02:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: VALLE

DATE/TIME OF TEST: 01/16/2016 03:33:57

Diagnostics Check		OK	03:27
Air Blank		0.000	03:27
Control Test		0.080	03:28
Air Blank		0.000	03:28
Subject Sample	# 1	0.124	03:29
Breath Volume		1.980	03:29
Air Blank		0.000	03:30
Air Blank		0.000	03:31
Subject Sample	# 2	0.124	03:32
Breath Volume		1.746	03:32
Air Blank		0.000	03:33
Control Test		0.079	03:33
Air Blank		0.000	03:33
Diagnostics Check		OK	03:33

Data Download Date/Time: 02/10/2016 08:55

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZRAK,JOSEPH,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: CRUZ,E
ARRESTING OFFICER AGENCY: Village of Pinecrest
LAST AGENCY INSPECTION: 01/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 02:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:06

SUBJECT'S LAST NAME: PICADO

DATE/TIME OF TEST: 01/21/2016 02:33:52

Diagnostics Check		OK	02:27
Air Blank		0.000	02:27
Control Test		0.080	02:27
Air Blank		0.000	02:28
Subject Sample	# 1	0.157	02:29
Breath Volume		3.132	02:29
Air Blank		0.000	02:29
Air Blank		0.000	02:31
Subject Sample	# 2	0.165	02:32
Breath Volume		3.851	02:32
Air Blank		0.000	02:32
Control Test		0.079	02:33
Air Blank		0.000	02:33
Diagnostics Check		OK	02:33

Data Download Date/Time: 02/10/2016 08:55

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZRAK,JOSEPH,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: AZRAK,JOSEPH
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 01/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 04:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:36

SUBJECT'S LAST NAME: SILBER

DATE/TIME OF TEST: 01/21/2016 04:59:45

Diagnostics Check		OK	04:56	
Air Blank		0.000	04:57	
Control Test		0.079	04:57	
Air Blank		0.000	04:58	
Subject Sample	# 1		04:58	Subject Test Refused
Breath Volume		0.000	04:58	
Air Blank		0.000	04:58	
Control Test		0.079	04:59	
Air Blank		0.000	04:59	
Diagnostics Check		OK	04:59	

Data Download Date/Time: 02/10/2016 08:55

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZRAK,JOSEPH,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: AZRAK,JOSEPH
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 01/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:33

SUBJECT'S LAST NAME: HALL JR

DATE/TIME OF TEST: 01/27/2016 02:02:34

Diagnostics Check		OK	01:56
Air Blank		0.000	01:56
Control Test		0.078	01:57
Air Blank		0.000	01:57
Subject Sample	# 1	0.362	01:58
Breath Volume		2.082	01:58
Air Blank		0.000	01:58
Air Blank		0.000	02:00
Subject Sample	# 2	0.350	02:00
Breath Volume		2.113	02:00
Air Blank		0.000	02:01
Control Test		0.078	02:01
Air Blank		0.000	02:02
Diagnostics Check		OK	02:02

Data Download Date/Time: 02/10/2016 08:55

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZRAK,JOSEPH,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: AZRAK,JOSEPH
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 01/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 00:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: COLLAZO FERNANDEZ

DATE/TIME OF TEST: 01/28/2016 00:36:56

Diagnostics Check		OK	00:30
Air Blank		0.000	00:30
Control Test		0.079	00:31
Air Blank		0.000	00:31
Subject Sample	# 1	0.194	00:32
Breath Volume		2.917	00:32
Air Blank		0.000	00:33
Air Blank		0.000	00:34
Subject Sample	# 2	0.203	00:35
Breath Volume		2.007	00:35
Air Blank		0.000	00:36
Control Test		0.079	00:36
Air Blank		0.000	00:36
Diagnostics Check		OK	00:36

Data Download Date/Time: 02/10/2016 08:55

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZRAK,JOSEPH,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: AZRAK,JOSEPH
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 01/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:36

SUBJECT'S LAST NAME: NUNEZ-CAPOTE

DATE/TIME OF TEST: 01/30/2016 01:08:29

Diagnostics Check		OK	00:59
Air Blank		0.000	01:00
Control Test		0.079	01:00
Air Blank		0.000	01:00
Subject Sample	# 1	0.213	01:03
Breath Volume		1.503	01:03
Air Blank		0.000	01:04
Air Blank		0.000	01:06
Subject Sample	# 2	0.209	01:06
Breath Volume		1.164	01:06
Air Blank		0.000	01:07
Control Test		0.079	01:07
Air Blank		0.000	01:08
Diagnostics Check		OK	01:08

Data Download Date/Time: 02/10/2016 08:55

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZRAK,JOSEPH,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: AZRAK,JOSEPH
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 01/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:36

SUBJECT'S LAST NAME: ALPIZA-GUZMAN

DATE/TIME OF TEST: 01/31/2016 03:05:01

Diagnostics Check		OK	02:58
Air Blank		0.000	02:59
Control Test		0.079	02:59
Air Blank		0.000	03:00
Subject Sample	# 1	0.229	03:00
Breath Volume		1.488	03:00
Air Blank		0.000	03:01
Air Blank		0.000	03:02
Subject Sample	# 2	0.220	03:03
Breath Volume		3.375	03:03
Air Blank		0.000	03:04
Control Test		0.079	03:04
Air Blank		0.000	03:04
Diagnostics Check		OK	03:05

Data Download Date/Time: 02/10/2016 08:55

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005810

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KELLNER,MICHAEL,A
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: HARRISON,NICHOLAS
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 12/15/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:47

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 01/03/2016 02:17:19

Diagnostics Check		OK	02:10
Air Blank		0.000	02:11
Control Test		0.078	02:11
Air Blank		0.000	02:11
Subject Sample	# 1	0.200	02:12
Breath Volume		2.191	02:12
Air Blank		0.000	02:13
Air Blank		0.000	02:15
Subject Sample	# 2	0.182	02:15
Breath Volume		2.296	02:15
Air Blank		0.000	02:16
Control Test		0.078	02:16
Air Blank		0.000	02:17
Diagnostics Check		OK	02:17

Data Download Date/Time: 01/25/2016 08:35

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005810

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KELLNER,MICHAEL,A
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: KELLNER,MICHAEL
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 12/15/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: BROCK III

DATE/TIME OF TEST: 01/13/2016 02:04:47

Diagnostics Check		OK	01:58
Air Blank		0.000	01:58
Control Test		0.080	01:58
Air Blank		0.000	01:59
Subject Sample	# 1	0.170	02:00
Breath Volume		2.050	02:00
Air Blank		0.000	02:00
Air Blank		0.000	02:02
Subject Sample	# 2	0.173	02:03
Breath Volume		1.742	02:03
Air Blank		0.000	02:03
Control Test		0.080	02:04
Air Blank		0.000	02:04
Diagnostics Check		OK	02:04

Data Download Date/Time: 01/25/2016 08:35

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005810

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KELLNER,MICHAEL,A
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: LANDRY,CHAD
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 12/15/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:04

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 01/17/2016 05:44:33

Diagnostics Check		OK	05:26	
Air Blank		0.000	05:27	
Control Test		0.078	05:27	
Air Blank		0.000	05:28	
Subject Sample	# 1	0.170	05:31	Volume Not Met
Breath Volume		0.390	05:31	
Air Blank		0.000	05:32	
Air Blank		0.000	05:33	
Subject Sample	# 2	0.217	05:36	
Breath Volume		1.566	05:36	
Air Blank		0.000	05:37	
Air Blank		0.000	05:39	
Subject Sample	# 3	0.218	05:42	Volume Not Met
Breath Volume		0.496	05:42	
Air Blank		0.000	05:43	
Control Test		0.072	05:43	Control Outside Tolerance
Air Blank		0.000	05:44	

Data Download Date/Time: 01/25/2016 08:35

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005810

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KELLNER,MICHAEL,A
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: LANDRY,CHAD
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 12/15/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:46

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 01/17/2016 06:10:11

Diagnostics Check	OK	06:08	
Air Blank	0.000	06:09	
Control Test	0.074	06:09	Control Outside Tolerance
Air Blank	0.000	06:10	

Data Download Date/Time: 01/25/2016 08:35

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005935

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLOW,CHRISTOPHER,A
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: BLOW,CHRISTOPHER
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 01/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 04:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: GTONZALEZ JR

DATE/TIME OF TEST: 01/23/2016 04:24:20

Diagnostics Check		OK	04:17
Air Blank		0.000	04:17
Control Test		0.082	04:18
Air Blank		0.000	04:18
Subject Sample	# 1	0.175	04:19
Breath Volume		2.445	04:19
Air Blank		0.000	04:19
Air Blank		0.000	04:21
Subject Sample	# 2	0.169	04:22
Breath Volume		2.117	04:22
Air Blank		0.000	04:23
Control Test		0.080	04:23
Air Blank		0.000	04:24
Diagnostics Check		OK	04:24

Data Download Date/Time: 02/08/2016 15:05

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005935

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLOW,CHRISTOPHER,A
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: BLOW,CHRISTOPHER
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 01/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 20:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:03

SUBJECT'S LAST NAME: VELEZ-AMADOR

DATE/TIME OF TEST: 01/24/2016 21:11:44

Diagnostics Check		OK	21:04
Air Blank		0.000	21:05
Control Test		0.084	21:05
Air Blank		0.000	21:06
Subject Sample	# 1	0.000	21:06
Breath Volume		1.394	21:06
Air Blank		0.000	21:07
Air Blank		0.000	21:09
Subject Sample	# 2	0.000	21:10
Breath Volume		2.664	21:10
Air Blank		0.000	21:10
Control Test		0.081	21:11
Air Blank		0.000	21:11
Diagnostics Check		OK	21:11

Data Download Date/Time: 02/08/2016 15:05

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005935

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLOW,CHRISTOPHER,A
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: BLOW,CHRISTOPHER
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 01/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 21:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: KING

DATE/TIME OF TEST: 01/28/2016 22:00:28

Diagnostics Check		OK	21:53
Air Blank		0.000	21:53
Control Test		0.082	21:54
Air Blank		0.000	21:54
Subject Sample	# 1	0.153	21:55
Breath Volume		1.152	21:55
Air Blank		0.000	21:55
Air Blank		0.000	21:57
Subject Sample	# 2	0.141	21:58
Breath Volume		1.492	21:58
Air Blank		0.000	21:59
Control Test		0.080	21:59
Air Blank		0.000	22:00
Diagnostics Check		OK	22:00

Data Download Date/Time: 02/08/2016 15:05

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005936

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PUFAHL,G,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: PUFAHL,G
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 12/28/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28814080A3
DRY GAS EXPIRY DATE: 11/05/2016
VIOLATION CODE: DUI
ARREST TIME: 01:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:37

SUBJECT'S LAST NAME: REYNOLDS

DATE/TIME OF TEST: 01/02/2016 02:06:26

Diagnostics Check		OK	01:57
Air Blank		0.000	01:58
Control Test		0.079	01:58
Air Blank		0.000	01:58
Subject Sample	# 1	0.158	02:00
Breath Volume		1.542	02:00
Air Blank		0.000	02:01
Air Blank		0.000	02:03
Subject Sample	# 2	0.154	02:04
Breath Volume		1.855	02:04
Air Blank		0.000	02:05
Control Test		0.079	02:05
Air Blank		0.000	02:06
Diagnostics Check		OK	02:06

Data Download Date/Time: 01/27/2016 10:18

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005936

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PUFAHL,G,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: PUFAHL,G
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 12/28/2015

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 28814080A3
DRY GAS EXPIRY DATE: 11/05/2016
VIOLATION CODE: DUI
ARREST TIME: 23:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:03

SUBJECT'S LAST NAME: OCONNOR

DATE/TIME OF TEST: 01/08/2016 23:32:15

Diagnostics Check		OK	23:23
Air Blank		0.000	23:24
Control Test		0.080	23:24
Air Blank		0.000	23:24
Subject Sample	# 1	0.155	23:26
Breath Volume		1.406	23:26
Air Blank		0.000	23:27
Air Blank		0.000	23:29
Subject Sample	# 2	0.155	23:30
Breath Volume		1.378	23:30
Air Blank		0.000	23:31
Control Test		0.078	23:31
Air Blank		0.000	23:32
Diagnostics Check		OK	23:32

Data Download Date/Time: 01/27/2016 10:18

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005936

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PUFAHL,G,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: PUFAHL,G
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 12/28/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28814080A3
DRY GAS EXPIRY DATE: 11/05/2016
VIOLATION CODE: DUI
ARREST TIME: 18:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:06

SUBJECT'S LAST NAME: EMMOTT

DATE/TIME OF TEST: 01/15/2016 18:35:37

Diagnostics Check		OK	18:26
Air Blank		0.000	18:26
Control Test		0.081	18:27
Air Blank		0.000	18:27
Subject Sample	# 1	0.269	18:28
Breath Volume		1.664	18:28
Air Blank		0.000	18:29
Air Blank		0.000	18:31
Subject Sample	# 2	0.262	18:33
Breath Volume		1.230	18:33
Air Blank		0.000	18:34
Control Test		0.081	18:34
Air Blank		0.000	18:35
Diagnostics Check		OK	18:35

Data Download Date/Time: 01/27/2016 10:18

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005936

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PUFAHL,G,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: ADAMS,J
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 12/28/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28814080A3
DRY GAS EXPIRY DATE: 11/05/2016
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: DUDA

DATE/TIME OF TEST: 01/22/2016 02:07:55

Diagnostics Check		OK	02:01
Air Blank		0.000	02:01
Control Test		0.081	02:01
Air Blank		0.000	02:02
Subject Sample	# 1	0.196	02:02
Breath Volume		1.382	02:02
Air Blank		0.000	02:03
Air Blank		0.000	02:05
Subject Sample	# 2	0.192	02:06
Breath Volume		2.097	02:06
Air Blank		0.000	02:06
Control Test		0.079	02:07
Air Blank		0.000	02:07
Diagnostics Check		OK	02:07

Data Download Date/Time: 01/27/2016 10:18

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005936

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PUFAHL,G,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: FLINT,C
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 12/28/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28814080A3
DRY GAS EXPIRY DATE: 11/05/2016
VIOLATION CODE: DUI
ARREST TIME: 20:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:36

SUBJECT'S LAST NAME: KHAN

DATE/TIME OF TEST: 01/23/2016 21:03:35

Diagnostics Check		OK	20:57
Air Blank		0.000	20:57
Control Test		0.081	20:57
Air Blank		0.000	20:58
Subject Sample	# 1	0.051	20:58
Breath Volume		3.574	20:58
Air Blank		0.000	20:59
Air Blank		0.000	21:01
Subject Sample	# 2	0.051	21:02
Breath Volume		4.015	21:02
Air Blank		0.000	21:02
Control Test		0.079	21:02
Air Blank		0.000	21:03
Diagnostics Check		OK	21:03

Data Download Date/Time: 01/27/2016 10:18

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005936

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PUFAHL,G,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: PUFAHL,G
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 01/27/2016

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 28814080A3
DRY GAS EXPIRY DATE: 11/05/2016
VIOLATION CODE: DUI
ARREST TIME: 20:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:01

SUBJECT'S LAST NAME: GODWIN

DATE/TIME OF TEST: 01/29/2016 21:27:34

Diagnostics Check		OK	21:21
Air Blank		0.000	21:22
Control Test		0.081	21:22
Air Blank		0.000	21:22
Subject Sample	# 1	0.000	21:23
Breath Volume		2.292	21:23
Air Blank		0.000	21:23
Air Blank		0.000	21:25
Subject Sample	# 2	0.000	21:26
Breath Volume		1.980	21:26
Air Blank		0.000	21:26
Control Test		0.080	21:26
Air Blank		0.000	21:27
Diagnostics Check		OK	21:27

Data Download Date/Time: 02/23/2016 08:24

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MERCIER,DANIEL
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13314080A3
DRY GAS EXPIRY DATE: 06/05/2016
VIOLATION CODE: DUI
ARREST TIME: 21:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:35

SUBJECT'S LAST NAME: JERONIMO

DATE/TIME OF TEST: 01/08/2016 22:13:48

Diagnostics Check		OK	22:07
Air Blank		0.000	22:07
Control Test		0.081	22:07
Air Blank		0.000	22:08
Subject Sample	# 1	0.079	22:09
Breath Volume		1.800	22:09
Air Blank		0.000	22:09
Air Blank		0.000	22:11
Subject Sample	# 2	0.084	22:12
Breath Volume		2.273	22:12
Air Blank		0.000	22:12
Control Test		0.080	22:13
Air Blank		0.000	22:13
Diagnostics Check		OK	22:13

Data Download Date/Time: 01/11/2016 13:22

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: GRUEBEL,KARL
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13314080A3
DRY GAS EXPIRY DATE: 06/05/2016
VIOLATION CODE: DUI
ARREST TIME: 00:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:37

SUBJECT'S LAST NAME: GOMOLL

DATE/TIME OF TEST: 01/09/2016 02:11:07

Diagnostics Check		OK	02:04
Air Blank		0.000	02:05
Control Test		0.080	02:05
Air Blank		0.000	02:05
Subject Sample	# 1	0.123	02:06
Breath Volume		1.878	02:06
Air Blank		0.000	02:07
Air Blank		0.000	02:09
Subject Sample	# 2	0.123	02:09
Breath Volume		1.687	02:09
Air Blank		0.000	02:10
Control Test		0.080	02:10
Air Blank		0.000	02:11
Diagnostics Check		OK	02:11

Data Download Date/Time: 01/11/2016 13:22

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MERCIER,DANIEL
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 01/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13314080A3
DRY GAS EXPIRY DATE: 06/05/2016
VIOLATION CODE: DUI
ARREST TIME: 22:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: BIRKELAND

DATE/TIME OF TEST: 01/23/2016 23:09:55

Diagnostics Check		OK	23:03
Air Blank		0.000	23:03
Control Test		0.081	23:04
Air Blank		0.000	23:04
Subject Sample	# 1	0.000	23:05
Breath Volume		1.480	23:05
Air Blank		0.000	23:06
Air Blank		0.000	23:08
Subject Sample	# 2	0.000	23:08
Breath Volume		2.035	23:08
Air Blank		0.000	23:09
Control Test		0.081	23:09
Air Blank		0.000	23:09
Diagnostics Check		OK	23:09

Data Download Date/Time: 02/17/2016 11:58

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006085

INSTRUMENT REGISTERED TO: US COAST GUARD MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHIRLEY,TREVOR,G
BREATH TEST OPERATOR AGENCY: USCG Miami
ARRESTING OFFICER: CARPER,SEAN
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 12/14/2015

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 09014080A1
DRY GAS EXPIRY DATE: 05/01/2016
VIOLATION CODE: BUI
ARREST TIME: 21:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:44

SUBJECT'S LAST NAME: SUDA

DATE/TIME OF TEST: 01/24/2016 23:00:16

Diagnostics Check	OK	22:59	
Air Blank	0.000	22:59	
Control Test	0.059	22:59	Control Outside Tolerance
Air Blank		23:00	RFI Detect

Data Download Date/Time: 01/26/2016 10:29

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006085

INSTRUMENT REGISTERED TO: US COAST GUARD MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHIRLEY,TREVOR,G
BREATH TEST OPERATOR AGENCY: USCG Miami
ARRESTING OFFICER: CARPER,SEAN
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 12/14/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09014080A1
DRY GAS EXPIRY DATE: 05/01/2016
VIOLATION CODE: BUI
ARREST TIME: 23:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:44

SUBJECT'S LAST NAME: SUDA

DATE/TIME OF TEST: 01/24/2016 23:23:43

Diagnostics Check		OK	23:20	
Air Blank		0.000	23:20	
Control Test		0.080	23:21	
Air Blank		0.000	23:21	
Subject Sample	# 1		23:22	Subject Test Refused
Breath Volume		0.000	23:22	
Air Blank		0.000	23:22	
Control Test		0.079	23:23	
Air Blank		0.000	23:23	
Diagnostics Check		OK	23:23	

Data Download Date/Time: 01/26/2016 10:29

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUN,DAVID,
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: YUN,DAVID
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A2
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:52

SUBJECT'S LAST NAME: CARTY

DATE/TIME OF TEST: 01/01/2016 04:25:18

Diagnostics Check		OK	04:13	
Air Blank		0.000	04:14	
Control Test		0.079	04:14	
Air Blank		0.000	04:14	
Subject Sample	# 1	0.000	04:18	Volume Not Met
Breath Volume		0.015	04:18	
Air Blank		0.000	04:18	
Air Blank		0.000	04:20	
Subject Sample	# 2	0.138	04:23	Volume Not Met
Breath Volume		0.414	04:23	
Air Blank		0.000	04:24	
Control Test		0.078	04:24	
Air Blank		0.000	04:25	
Diagnostics Check		OK	04:25	

Data Download Date/Time: 01/27/2016 13:14

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUN,DAVID,
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: YUN,DAVID
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A2
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:52

SUBJECT'S LAST NAME: CARTY

DATE/TIME OF TEST: 01/01/2016 04:35:21

Diagnostics Check		OK	04:28
Air Blank		0.000	04:28
Control Test		0.077	04:29
Air Blank		0.000	04:29
Subject Sample	# 1	0.157	04:30
Breath Volume		1.785	04:30
Air Blank		0.000	04:31
Air Blank		0.000	04:33
Subject Sample	# 2	0.153	04:33
Breath Volume		1.765	04:33
Air Blank		0.000	04:34
Control Test		0.075	04:34
Air Blank		0.000	04:35
Diagnostics Check		OK	04:35

Data Download Date/Time: 01/27/2016 13:14

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAKER,BRADFORD,B
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: BAKER,BRADFORD
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A2
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: MORALES

DATE/TIME OF TEST: 01/02/2016 03:15:11

Diagnostics Check		OK	03:08
Air Blank		0.000	03:08
Control Test		0.078	03:08
Air Blank		0.000	03:09
Subject Sample	# 1	0.210	03:09
Breath Volume		2.496	03:09
Air Blank		0.000	03:10
Air Blank		0.000	03:12
Subject Sample	# 2	0.219	03:13
Breath Volume		2.730	03:13
Air Blank		0.000	03:14
Control Test		0.075	03:14
Air Blank		0.000	03:15
Diagnostics Check		OK	03:15

Data Download Date/Time: 01/27/2016 13:14

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POLIZZI,CHRISTINE,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: POLIZZI,CHRISTINE
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A2
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 18:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:00

SUBJECT'S LAST NAME: WAMSLEY

DATE/TIME OF TEST: 01/02/2016 19:27:15

Diagnostics Check		OK	19:20
Air Blank		0.000	19:21
Control Test		0.079	19:21
Air Blank		0.000	19:22
Subject Sample	# 1	0.045	19:22
Breath Volume		2.683	19:22
Air Blank		0.000	19:23
Air Blank		0.000	19:25
Subject Sample	# 2	0.045	19:25
Breath Volume		2.542	19:25
Air Blank		0.000	19:26
Control Test		0.079	19:26
Air Blank		0.000	19:27
Diagnostics Check		OK	19:27

Data Download Date/Time: 01/27/2016 13:14

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUN,DAVID,
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: YUN,DAVID
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A2
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:48

SUBJECT'S LAST NAME: SNOOK

DATE/TIME OF TEST: 01/09/2016 03:15:23

Diagnostics Check		OK	03:08
Air Blank		0.000	03:09
Control Test		0.079	03:09
Air Blank		0.000	03:10
Subject Sample	# 1	0.167	03:10
Breath Volume		2.468	03:10
Air Blank		0.000	03:11
Air Blank		0.000	03:13
Subject Sample	# 2	0.169	03:13
Breath Volume		2.886	03:13
Air Blank		0.000	03:14
Control Test		0.079	03:14
Air Blank		0.000	03:15
Diagnostics Check		OK	03:15

Data Download Date/Time: 01/27/2016 13:14

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KRECIC,COREY,J
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: KRECIC,COREY
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A2
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 15:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:21

SUBJECT'S LAST NAME: KARRAS

DATE/TIME OF TEST: 01/09/2016 16:57:30

Diagnostics Check		OK	16:51
Air Blank		0.000	16:51
Control Test		0.079	16:51
Air Blank		0.000	16:52
Subject Sample	# 1	0.170	16:52
Breath Volume		1.796	16:52
Air Blank		0.000	16:53
Air Blank		0.000	16:55
Subject Sample	# 2	0.178	16:55
Breath Volume		1.406	16:55
Air Blank		0.000	16:56
Control Test		0.077	16:56
Air Blank		0.000	16:57
Diagnostics Check		OK	16:57

Data Download Date/Time: 01/27/2016 13:14

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POLIZZI,CHRISTINE,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: POLIZZI,CHRISTINE
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A2
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 19:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:05

SUBJECT'S LAST NAME: GASPAR

DATE/TIME OF TEST: 01/09/2016 20:34:27

Diagnostics Check		OK	20:27
Air Blank		0.000	20:27
Control Test		0.079	20:27
Air Blank		0.000	20:28
Subject Sample	# 1	0.132	20:29
Breath Volume		2.054	20:29
Air Blank		0.000	20:29
Air Blank		0.000	20:31
Subject Sample	# 2	0.124	20:32
Breath Volume		2.164	20:32
Air Blank		0.000	20:33
Control Test		0.079	20:33
Air Blank		0.000	20:34
Diagnostics Check		OK	20:34

Data Download Date/Time: 01/27/2016 13:14

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PORCELLI,ANDREW,
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: WISE,EUGENE
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A2
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 15:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:36

SUBJECT'S LAST NAME: HUBBARD

DATE/TIME OF TEST: 01/11/2016 16:38:28

Diagnostics Check		OK	16:32
Air Blank		0.000	16:32
Control Test		0.078	16:33
Air Blank		0.000	16:33
Subject Sample	# 1	0.000	16:34
Breath Volume		3.109	16:34
Air Blank		0.000	16:34
Air Blank		0.000	16:36
Subject Sample	# 2	0.000	16:37
Breath Volume		3.078	16:37
Air Blank		0.000	16:37
Control Test		0.076	16:37
Air Blank		0.000	16:38
Diagnostics Check		OK	16:38

Data Download Date/Time: 01/27/2016 13:14

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PECCI,Z,A
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: PECCI,Z
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A2
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: NYMAN

DATE/TIME OF TEST: 01/12/2016 02:49:50

Diagnostics Check		OK	02:40	
Air Blank		0.000	02:41	
Control Test		0.079	02:41	
Air Blank		0.000	02:41	
Subject Sample	# 1		02:45	No Sample Provided
Breath Volume		0.000	02:45	
Air Blank		0.000	02:45	
Air Blank		0.000	02:47	
Subject Sample	# 2		02:48	Subject Test Refused
Breath Volume		0.000	02:48	
Air Blank		0.000	02:48	
Control Test		0.079	02:49	
Air Blank		0.000	02:49	
Diagnostics Check		OK	02:49	

Data Download Date/Time: 01/27/2016 13:14

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A2
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: PACIFICI

DATE/TIME OF TEST: 01/12/2016 03:21:54

Diagnostics Check		OK	03:14
Air Blank		0.000	03:14
Control Test		0.079	03:15
Air Blank		0.000	03:15
Subject Sample	# 1	0.111	03:16
Breath Volume		2.789	03:16
Air Blank		0.000	03:16
Air Blank		0.000	03:18
Subject Sample	# 2	0.105	03:20
Breath Volume		2.140	03:20
Air Blank		0.000	03:20
Control Test		0.079	03:21
Air Blank		0.000	03:21
Diagnostics Check		OK	03:21

Data Download Date/Time: 01/27/2016 13:14

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A2
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:24

SUBJECT'S LAST NAME: IVORY

DATE/TIME OF TEST: 01/13/2016 01:01:09

Diagnostics Check		OK	00:54
Air Blank		0.000	00:55
Control Test		0.080	00:55
Air Blank		0.000	00:56
Subject Sample	# 1	0.084	00:56
Breath Volume		2.507	00:56
Air Blank		0.000	00:57
Air Blank		0.000	00:59
Subject Sample	# 2	0.082	00:59
Breath Volume		2.656	00:59
Air Blank		0.000	01:00
Control Test		0.079	01:00
Air Blank		0.000	01:01
Diagnostics Check		OK	01:01

Data Download Date/Time: 01/27/2016 13:14

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CUTY,J,K
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: CUTY,J
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A2
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 14:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:39

SUBJECT'S LAST NAME: CLOUTIER

DATE/TIME OF TEST: 01/15/2016 15:08:37

Diagnostics Check		OK	15:02
Air Blank		0.000	15:02
Control Test		0.078	15:03
Air Blank		0.000	15:03
Subject Sample	# 1	0.235	15:04
Breath Volume		2.328	15:04
Air Blank		0.000	15:04
Air Blank		0.000	15:06
Subject Sample	# 2	0.239	15:07
Breath Volume		2.960	15:07
Air Blank		0.000	15:07
Control Test		0.078	15:08
Air Blank		0.000	15:08
Diagnostics Check		OK	15:08

Data Download Date/Time: 01/27/2016 13:14

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POLIZZI,CHRISTINE,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: POLIZZI,CHRISTINE
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A2
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: DISANGRO

DATE/TIME OF TEST: 01/17/2016 00:43:28

Diagnostics Check		OK	00:37
Air Blank		0.000	00:37
Control Test		0.079	00:37
Air Blank		0.000	00:38
Subject Sample	# 1	0.135	00:38
Breath Volume		2.023	00:38
Air Blank		0.000	00:39
Air Blank		0.000	00:41
Subject Sample	# 2	0.132	00:41
Breath Volume		1.937	00:41
Air Blank		0.000	00:42
Control Test		0.079	00:42
Air Blank		0.000	00:43
Diagnostics Check		OK	00:43

Data Download Date/Time: 01/27/2016 13:14

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEELE,MARK,T
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: STEELE,MARK
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A2
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:53

SUBJECT'S LAST NAME: UNKNOWN

DATE/TIME OF TEST: 01/17/2016 01:21:38

Diagnostics Check		OK	01:13
Air Blank		0.000	01:14
Control Test		0.079	01:14
Air Blank		0.000	01:15
Subject Sample	# 1	0.049	01:15
Breath Volume		2.285	01:15
Air Blank		0.000	01:16
Air Blank		0.000	01:18
Subject Sample	# 2	0.045	01:19
Breath Volume		1.503	01:19
Air Blank		0.000	01:20
Control Test		0.079	01:21
Air Blank		0.000	01:21
Diagnostics Check		OK	01:21

Data Download Date/Time: 01/27/2016 13:14

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PORCELLI,ANDREW,
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: PORCELLI,ANDREW
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A2
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 13:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:09

SUBJECT'S LAST NAME: MIGUEL

DATE/TIME OF TEST: 01/17/2016 13:42:19

Diagnostics Check		OK	13:37	
Air Blank		0.000	13:37	
Control Test		0.079	13:38	
Air Blank		0.000	13:38	
Subject Sample	# 1		13:41	Sequence Aborted
Breath Volume		0.000	13:41	
Air Blank		0.000	13:42	

Data Download Date/Time: 01/27/2016 13:14

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUN,DAVID,
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: YUN,DAVID
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A2
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 17:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:55

SUBJECT'S LAST NAME: VOLPE

DATE/TIME OF TEST: 01/17/2016 18:36:15

Diagnostics Check		OK	18:25
Air Blank		0.000	18:25
Control Test		0.079	18:26
Air Blank		0.000	18:26
Subject Sample	# 1	0.211	18:27
Breath Volume		1.652	18:27
Air Blank		0.000	18:28
Air Blank		0.000	18:30
Subject Sample	# 2	0.188	18:31
Breath Volume		1.832	18:31
Air Blank		0.000	18:31
Air Blank		0.000	18:33
Subject Sample	# 3	0.191	18:34
Breath Volume		2.062	18:34
Air Blank		0.000	18:35
Control Test		0.077	18:35
Air Blank		0.000	18:36
Diagnostics Check		OK	18:36

Data Download Date/Time: 01/27/2016 13:14

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAKER,BRADFORD,B
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: BAKER,BRADFORD
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A2
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 19:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:19

SUBJECT'S LAST NAME: PETTIT II

DATE/TIME OF TEST: 01/19/2016 20:55:27

Diagnostics Check		OK	20:48
Air Blank		0.000	20:49
Control Test		0.079	20:49
Air Blank		0.000	20:50
Subject Sample	# 1	0.148	20:50
Breath Volume		3.925	20:50
Air Blank		0.000	20:51
Air Blank		0.000	20:53
Subject Sample	# 2	0.144	20:53
Breath Volume		3.343	20:53
Air Blank		0.000	20:54
Control Test		0.076	20:54
Air Blank		0.000	20:55
Diagnostics Check		OK	20:55

Data Download Date/Time: 01/27/2016 13:14

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POLIZZI,CHRISTINE,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: POLIZZI,CHRISTINE
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A2
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: TAX TZOC

DATE/TIME OF TEST: 01/22/2016 23:58:21

Diagnostics Check		OK	23:51
Air Blank		0.000	23:52
Control Test		0.079	23:52
Air Blank		0.000	23:53
Subject Sample	# 1	0.187	23:53
Breath Volume		3.105	23:53
Air Blank		0.000	23:54
Air Blank		0.000	23:56
Subject Sample	# 2	0.185	23:56
Breath Volume		3.667	23:56
Air Blank		0.000	23:57
Control Test		0.077	23:57
Air Blank		0.000	23:58
Diagnostics Check		OK	23:58

Data Download Date/Time: 01/27/2016 13:14

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAKER,BRADFORD,B
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: BAKER,BRADFORD
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A2
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:44

SUBJECT'S LAST NAME: SONTAY AJTUN

DATE/TIME OF TEST: 01/23/2016 01:17:21

Diagnostics Check		OK	01:08	
Air Blank		0.000	01:08	
Control Test		0.077	01:08	
Air Blank		0.000	01:09	
Subject Sample	# 1	0.223	01:10	
Breath Volume		1.367	01:10	
Air Blank		0.000	01:11	
Air Blank		0.000	01:12	
Subject Sample	# 2	0.214	01:15	Volume Not Met
Breath Volume		0.628	01:15	
Air Blank			01:16	Purge Fail
Air Blank		0.000	01:17	

Data Download Date/Time: 01/27/2016 13:14

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAKER,BRADFORD,B
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: BAKER,BRADFORD
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A2
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:44

SUBJECT'S LAST NAME: SONTAY AJTUN

DATE/TIME OF TEST: 01/23/2016 01:20:13

Diagnostics Check	OK	01:18	
Air Blank	0.000	01:19	
Control Test	0.074	01:19	Control Outside Tolerance
Air Blank	0.000	01:20	

Data Download Date/Time: 01/27/2016 13:14

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAKER,BRADFORD,B
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: BAKER,BRADFORD
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A2
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:44

SUBJECT'S LAST NAME: SONTAY AJTUN

DATE/TIME OF TEST: 01/23/2016 01:29:16

Diagnostics Check		OK	01:22
Air Blank		0.000	01:22
Control Test		0.077	01:23
Air Blank		0.000	01:23
Subject Sample	# 1	0.219	01:24
Breath Volume		1.828	01:24
Air Blank		0.000	01:25
Air Blank		0.000	01:26
Subject Sample	# 2	0.201	01:27
Breath Volume		1.664	01:27
Air Blank		0.000	01:28
Control Test		0.076	01:28
Air Blank		0.000	01:29
Diagnostics Check		OK	01:29

Data Download Date/Time: 01/27/2016 13:14

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POLIZZI,CHRISTINE,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: POLIZZI,CHRISTINE
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A2
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 17:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:06

SUBJECT'S LAST NAME: MUTH

DATE/TIME OF TEST: 01/23/2016 18:34:06

Diagnostics Check		OK	18:27
Air Blank		0.000	18:28
Control Test		0.080	18:28
Air Blank		0.000	18:28
Subject Sample	# 1	0.077	18:29
Breath Volume		2.113	18:29
Air Blank		0.000	18:30
Air Blank		0.000	18:31
Subject Sample	# 2	0.078	18:32
Breath Volume		1.867	18:32
Air Blank		0.000	18:33
Control Test		0.080	18:33
Air Blank		0.000	18:34
Diagnostics Check		OK	18:34

Data Download Date/Time: 01/27/2016 13:14

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: TILLET,JONPAUL
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A2
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:10

SUBJECT'S LAST NAME: RAYBORN II

DATE/TIME OF TEST: 01/23/2016 22:43:13

Diagnostics Check		OK	22:36
Air Blank		0.000	22:37
Control Test		0.079	22:37
Air Blank		0.000	22:38
Subject Sample	# 1	0.173	22:38
Breath Volume		1.535	22:38
Air Blank		0.000	22:39
Air Blank		0.000	22:41
Subject Sample	# 2	0.175	22:41
Breath Volume		2.164	22:41
Air Blank		0.000	22:42
Control Test		0.080	22:42
Air Blank		0.000	22:43
Diagnostics Check		OK	22:43

Data Download Date/Time: 01/27/2016 13:14

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACOBSON,DAVID,E
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: ANGERVIL,CLAUDY
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A2
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 18:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:58

SUBJECT'S LAST NAME: DE BAISE JR

DATE/TIME OF TEST: 01/24/2016 19:25:06

Diagnostics Check		OK	19:18
Air Blank		0.000	19:19
Control Test		0.080	19:19
Air Blank		0.000	19:20
Subject Sample	# 1	0.142	19:20
Breath Volume		3.316	19:20
Air Blank		0.000	19:21
Air Blank		0.000	19:22
Subject Sample	# 2	0.138	19:23
Breath Volume		3.039	19:23
Air Blank		0.000	19:24
Control Test		0.079	19:24
Air Blank		0.000	19:25
Diagnostics Check		OK	19:25

Data Download Date/Time: 01/27/2016 13:14

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GITTO,JUSTIN,J
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: GITTO,JUSTIN
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A2
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: MAGEE

DATE/TIME OF TEST: 01/26/2016 02:12:38

Diagnostics Check		OK	02:06
Air Blank		0.000	02:06
Control Test		0.079	02:06
Air Blank		0.000	02:07
Subject Sample	# 1	0.125	02:07
Breath Volume		2.308	02:07
Air Blank		0.000	02:08
Air Blank		0.000	02:10
Subject Sample	# 2	0.120	02:10
Breath Volume		1.640	02:10
Air Blank		0.000	02:11
Control Test		0.079	02:12
Air Blank		0.000	02:12
Diagnostics Check		OK	02:12

Data Download Date/Time: 01/27/2016 13:14

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 01/27/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A2
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: RECKAMP

DATE/TIME OF TEST: 01/28/2016 02:19:25

Diagnostics Check		OK	02:11
Air Blank		0.000	02:12
Control Test		0.078	02:12
Air Blank		0.000	02:13
Subject Sample	# 1	0.174	02:14
Breath Volume		1.867	02:14
Air Blank		0.000	02:14
Air Blank		0.000	02:16
Subject Sample	# 2	0.170	02:17
Breath Volume		1.730	02:17
Air Blank		0.000	02:18
Control Test		0.078	02:18
Air Blank		0.000	02:19
Diagnostics Check		OK	02:19

Data Download Date/Time: 02/17/2016 03:57

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEELE,MARK,T
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: STEELE,MARK
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 01/27/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A2
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:28

SUBJECT'S LAST NAME: COLDREN

DATE/TIME OF TEST: 01/28/2016 23:56:19

Diagnostics Check		OK	23:48
Air Blank		0.000	23:49
Control Test		0.078	23:49
Air Blank		0.000	23:50
Subject Sample	# 1	0.168	23:51
Breath Volume		2.531	23:51
Air Blank		0.000	23:51
Air Blank		0.000	23:53
Subject Sample	# 2	0.166	23:54
Breath Volume		2.066	23:54
Air Blank		0.000	23:55
Control Test		0.078	23:55
Air Blank		0.000	23:56
Diagnostics Check		OK	23:56

Data Download Date/Time: 02/17/2016 03:57

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAKER,BRADFORD,B
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: BAKER,BRADFORD
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 01/27/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A2
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: GONZALEZ MEJIA

DATE/TIME OF TEST: 01/29/2016 01:25:15

Diagnostics Check		OK	01:18
Air Blank		0.000	01:18
Control Test		0.078	01:19
Air Blank		0.000	01:19
Subject Sample	# 1	0.140	01:20
Breath Volume		1.519	01:20
Air Blank		0.000	01:21
Air Blank		0.000	01:22
Subject Sample	# 2	0.145	01:23
Breath Volume		1.562	01:23
Air Blank		0.000	01:24
Control Test		0.076	01:24
Air Blank		0.000	01:25
Diagnostics Check		OK	01:25

Data Download Date/Time: 02/17/2016 03:57

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICE,MICHAEL,D
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: ANGERVIL,CLAUDY
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 01/27/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A2
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: CASSIDY

DATE/TIME OF TEST: 01/29/2016 04:20:45

Diagnostics Check		OK	04:13
Air Blank		0.000	04:14
Control Test		0.078	04:14
Air Blank		0.000	04:15
Subject Sample	# 1	0.170	04:15
Breath Volume		2.753	04:15
Air Blank		0.000	04:16
Air Blank		0.000	04:18
Subject Sample	# 2	0.168	04:19
Breath Volume		2.488	04:19
Air Blank		0.000	04:19
Control Test		0.078	04:20
Air Blank		0.000	04:20
Diagnostics Check		OK	04:20

Data Download Date/Time: 02/17/2016 03:57

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POLIZZI,CHRISTINE,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: POLIZZI,CHRISTINE
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 01/27/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A2
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 18:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:13

SUBJECT'S LAST NAME: PEREZ-MASARIEGOS

DATE/TIME OF TEST: 01/29/2016 19:48:42

Diagnostics Check		OK	19:41
Air Blank		0.000	19:42
Control Test		0.079	19:42
Air Blank		0.000	19:43
Subject Sample	# 1	0.233	19:43
Breath Volume		2.558	19:43
Air Blank		0.000	19:44
Air Blank		0.000	19:46
Subject Sample	# 2	0.231	19:46
Breath Volume		2.605	19:46
Air Blank		0.000	19:47
Control Test		0.078	19:48
Air Blank		0.000	19:48
Diagnostics Check		OK	19:48

Data Download Date/Time: 02/17/2016 03:57

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POLIZZI,CHRISTINE,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: POLIZZI,CHRISTINE
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 01/27/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A2
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: MCQUEEN

DATE/TIME OF TEST: 01/29/2016 23:23:18

Diagnostics Check		OK	23:17	
Air Blank		0.000	23:18	
Control Test		0.079	23:18	
Air Blank		0.000	23:19	
Subject Sample	# 1		23:21	Subject Test Refused
Breath Volume		0.019	23:21	
Air Blank		0.000	23:22	
Control Test		0.079	23:22	
Air Blank		0.000	23:23	
Diagnostics Check		OK	23:23	

Data Download Date/Time: 02/17/2016 03:57

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GITTO,JUSTIN,J
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: GITTO,JUSTIN
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 01/27/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A2
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:36

SUBJECT'S LAST NAME: NOBREGAS

DATE/TIME OF TEST: 01/31/2016 04:04:16

Diagnostics Check		OK	03:57
Air Blank		0.000	03:58
Control Test		0.079	03:58
Air Blank		0.000	03:59
Subject Sample	# 1	0.086	03:59
Breath Volume		1.839	03:59
Air Blank		0.000	04:00
Air Blank		0.000	04:02
Subject Sample	# 2	0.091	04:02
Breath Volume		1.332	04:02
Air Blank		0.000	04:03
Control Test		0.078	04:03
Air Blank		0.000	04:04
Diagnostics Check		OK	04:04

Data Download Date/Time: 02/17/2016 03:57

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A2
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 19:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:41

SUBJECT'S LAST NAME: MEDINA-RAMIREZ

DATE/TIME OF TEST: 01/20/2016 20:09:40

Diagnostics Check		OK	20:03
Air Blank		0.000	20:03
Control Test		0.079	20:03
Air Blank		0.000	20:04
Subject Sample	# 1	0.011	20:05
Breath Volume		3.703	20:05
Air Blank		0.000	20:05
Air Blank		0.000	20:07
Subject Sample	# 2	0.008	20:08
Breath Volume		1.167	20:08
Air Blank		0.000	20:08
Control Test		0.080	20:09
Air Blank		0.000	20:09
Diagnostics Check		OK	20:09

Data Download Date/Time: 01/27/2016 13:01

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A2
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: BIONDO

DATE/TIME OF TEST: 01/24/2016 03:37:37

Diagnostics Check		OK	03:31
Air Blank		0.000	03:31
Control Test		0.080	03:31
Air Blank		0.000	03:32
Subject Sample	# 1	0.161	03:33
Breath Volume		4.867	03:33
Air Blank		0.000	03:33
Air Blank		0.000	03:35
Subject Sample	# 2	0.160	03:36
Breath Volume		4.433	03:36
Air Blank		0.000	03:36
Control Test		0.080	03:36
Air Blank		0.000	03:37
Diagnostics Check		OK	03:37

Data Download Date/Time: 01/27/2016 13:01

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,RAUL,
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: GONZALEZ,JESUS
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 12/23/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1673300
DRY GAS EXPIRY DATE: 04/28/2017
VIOLATION CODE: DUI
ARREST TIME: 03:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:09

SUBJECT'S LAST NAME: MARTINES

DATE/TIME OF TEST: 01/04/2016 04:42:05

Diagnostics Check		OK	04:33	
Air Blank		0.000	04:34	
Control Test		0.082	04:34	
Air Blank		0.000	04:34	
Subject Sample	# 1		04:38	No Sample Provided
Breath Volume		0.000	04:38	
Air Blank		0.000	04:38	
Air Blank		0.000	04:40	
Subject Sample	# 2		04:40	Subject Test Refused
Breath Volume		0.000	04:40	
Air Blank		0.000	04:41	
Control Test		0.081	04:41	
Air Blank		0.000	04:41	
Diagnostics Check		OK	04:42	

Data Download Date/Time: 01/28/2016 14:33

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REYES,GEORGE,
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: MONACO,LANCE
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 12/23/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1673300
DRY GAS EXPIRY DATE: 04/28/2017
VIOLATION CODE: DUI
ARREST TIME: 19:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:21

SUBJECT'S LAST NAME: PASCUAL-MARTIN

DATE/TIME OF TEST: 01/15/2016 19:55:48

Diagnostics Check		OK	19:49
Air Blank		0.000	19:49
Control Test		0.080	19:49
Air Blank		0.000	19:50
Subject Sample	# 1	0.117	19:51
Breath Volume		3.199	19:51
Air Blank		0.000	19:51
Air Blank		0.000	19:53
Subject Sample	# 2	0.114	19:54
Breath Volume		2.667	19:54
Air Blank		0.000	19:54
Control Test		0.080	19:55
Air Blank		0.000	19:55
Diagnostics Check		OK	19:55

Data Download Date/Time: 01/28/2016 14:33

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,JESUS,O
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: RODRIGUEZ,R
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 12/23/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1673300
DRY GAS EXPIRY DATE: 04/28/2017
VIOLATION CODE: DUI
ARREST TIME: 02:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:17

SUBJECT'S LAST NAME: OLVERA

DATE/TIME OF TEST: 01/19/2016 02:52:25

Diagnostics Check		OK	02:49	
Air Blank		0.000	02:50	
Control Test		0.082	02:50	
Air Blank		0.000	02:50	
Subject Sample	# 1		02:51	Subject Test Refused
Breath Volume		0.000	02:51	
Air Blank		0.000	02:51	
Control Test		0.082	02:51	
Air Blank		0.000	02:52	
Diagnostics Check		OK	02:52	

Data Download Date/Time: 01/28/2016 14:33

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEPHENSON,JENNIFER,L
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: GUZMAN,ENGELBERT
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 12/23/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1673300
DRY GAS EXPIRY DATE: 04/28/2017
VIOLATION CODE: DUI
ARREST TIME: 04:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:31

SUBJECT'S LAST NAME: JAUREGUI

DATE/TIME OF TEST: 01/23/2016 05:06:51

Diagnostics Check		OK	05:04	
Air Blank		0.000	05:04	
Control Test		0.079	05:04	
Air Blank		0.000	05:05	
Subject Sample	# 1		05:05	Subject Test Refused
Breath Volume		0.000	05:05	
Air Blank		0.000	05:05	
Control Test		0.080	05:06	
Air Blank		0.000	05:06	
Diagnostics Check		OK	05:06	

Data Download Date/Time: 01/28/2016 14:33

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REYES,GEORGE,
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: REYES,GEORGE
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 12/23/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1673300
DRY GAS EXPIRY DATE: 04/28/2017
VIOLATION CODE: DUI
ARREST TIME: 23:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:52

SUBJECT'S LAST NAME: VICENTE-HERNANDEZ

DATE/TIME OF TEST: 01/27/2016 00:33:57

Diagnostics Check		OK	00:27
Air Blank		0.000	00:27
Control Test		0.081	00:27
Air Blank		0.000	00:28
Subject Sample	# 1	0.235	00:29
Breath Volume		2.218	00:29
Air Blank		0.000	00:30
Air Blank		0.000	00:31
Subject Sample	# 2	0.239	00:32
Breath Volume		1.949	00:32
Air Blank		0.000	00:33
Control Test		0.080	00:33
Air Blank		0.000	00:33
Diagnostics Check		OK	00:33

Data Download Date/Time: 01/28/2016 14:33

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SURMAN,THOMAS,A
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: MILLER,TODD
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 12/23/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1673300
DRY GAS EXPIRY DATE: 04/28/2017
VIOLATION CODE: DUI
ARREST TIME: 10:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:40

SUBJECT'S LAST NAME: GIBSON

DATE/TIME OF TEST: 01/28/2016 11:23:28

Diagnostics Check		OK	11:12
Air Blank		0.000	11:13
Control Test		0.081	11:13
Air Blank		0.000	11:14
Subject Sample	# 1	0.074	11:17
Breath Volume		1.316	11:17
Air Blank		0.000	11:17
Air Blank		0.000	11:19
Subject Sample	# 2	0.074	11:21
Breath Volume		1.316	11:21
Air Blank		0.000	11:22
Control Test		0.080	11:22
Air Blank		0.000	11:23
Diagnostics Check		OK	11:23

Data Download Date/Time: 01/28/2016 14:33

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006214

INSTRUMENT REGISTERED TO: OCALA POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOEDICKE,WILLIAM,R
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: MOOREHEAD,DAVID
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 01/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:26

SUBJECT'S LAST NAME: SPRIGGS

DATE/TIME OF TEST: 01/09/2016 00:53:41

Diagnostics Check		OK	00:47
Air Blank		0.000	00:47
Control Test		0.078	00:48
Air Blank		0.000	00:48
Subject Sample	# 1	0.170	00:49
Breath Volume		1.406	00:49
Air Blank		0.000	00:49
Air Blank		0.000	00:51
Subject Sample	# 2	0.175	00:52
Breath Volume		1.988	00:52
Air Blank		0.000	00:52
Control Test		0.077	00:53
Air Blank		0.000	00:53
Diagnostics Check		OK	00:53

Data Download Date/Time: 02/25/2016 00:14

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006214

INSTRUMENT REGISTERED TO: OCALA POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOEDICKE,WILLIAM,R
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: JOEDICKE,WILLIAM
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 01/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:44

SUBJECT'S LAST NAME: LANGFORD

DATE/TIME OF TEST: 01/09/2016 02:10:23

Diagnostics Check		OK	02:04
Air Blank		0.000	02:04
Control Test		0.078	02:05
Air Blank		0.000	02:05
Subject Sample	# 1	0.123	02:05
Breath Volume		3.199	02:05
Air Blank		0.000	02:06
Air Blank		0.000	02:08
Subject Sample	# 2	0.119	02:08
Breath Volume		3.507	02:08
Air Blank		0.000	02:09
Control Test		0.077	02:09
Air Blank		0.000	02:10
Diagnostics Check		OK	02:10

Data Download Date/Time: 02/25/2016 00:14

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006214

INSTRUMENT REGISTERED TO: OCALA POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOEDICKE,WILLIAM,R
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: MOOREHEAD,DAVID
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 01/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:45

SUBJECT'S LAST NAME: BEEMER

DATE/TIME OF TEST: 01/14/2016 21:15:24

Diagnostics Check		OK	21:09
Air Blank		0.000	21:09
Control Test		0.080	21:09
Air Blank		0.000	21:10
Subject Sample	# 1	0.120	21:10
Breath Volume		3.800	21:10
Air Blank		0.000	21:11
Air Blank		0.000	21:13
Subject Sample	# 2	0.118	21:13
Breath Volume		3.574	21:13
Air Blank		0.000	21:14
Control Test		0.077	21:14
Air Blank		0.000	21:15
Diagnostics Check		OK	21:15

Data Download Date/Time: 02/25/2016 00:14

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006214

INSTRUMENT REGISTERED TO: OCALA POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOEDICKE,WILLIAM,R
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: JOEDICKE,WILLIAM
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 01/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: RAMDOOIAR

DATE/TIME OF TEST: 01/15/2016 23:23:04

Diagnostics Check		OK	23:16
Air Blank		0.000	23:17
Control Test		0.079	23:17
Air Blank		0.000	23:18
Subject Sample	# 1	0.000	23:18
Breath Volume		2.035	23:18
Air Blank		0.000	23:19
Air Blank		0.000	23:21
Subject Sample	# 2	0.000	23:21
Breath Volume		1.820	23:21
Air Blank		0.000	23:22
Control Test		0.078	23:22
Air Blank		0.000	23:22
Diagnostics Check		OK	23:23

Data Download Date/Time: 02/25/2016 00:14

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006214

INSTRUMENT REGISTERED TO: OCALA POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOEDICKE,WILLIAM,R
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: JOEDICKE,WILLIAM
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 01/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 19:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:17

SUBJECT'S LAST NAME: WELCH

DATE/TIME OF TEST: 01/16/2016 19:58:51

Diagnostics Check		OK	19:47	
Air Blank		0.000	19:47	
Control Test		0.078	19:48	
Air Blank		0.000	19:48	
Subject Sample	# 1		19:51	No Sample Provided
Breath Volume		0.000	19:51	
Air Blank		0.000	19:52	
Air Blank		0.000	19:54	
Subject Sample	# 2		19:57	No Sample Provided
Breath Volume		0.000	19:57	
Air Blank		0.000	19:57	
Control Test		0.078	19:58	
Air Blank		0.000	19:58	
Diagnostics Check		OK	19:58	

Data Download Date/Time: 02/25/2016 00:14

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006214

INSTRUMENT REGISTERED TO: OCALA POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOEDICKE,WILLIAM,R
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: JOEDICKE,WILLIAM
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 01/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: HINKLE

DATE/TIME OF TEST: 01/17/2016 00:22:11

Diagnostics Check		OK	00:15
Air Blank		0.000	00:16
Control Test		0.078	00:16
Air Blank		0.000	00:17
Subject Sample	# 1	0.000	00:17
Breath Volume		3.398	00:17
Air Blank		0.000	00:18
Air Blank		0.000	00:20
Subject Sample	# 2	0.000	00:20
Breath Volume		3.769	00:20
Air Blank		0.000	00:21
Control Test		0.077	00:21
Air Blank		0.000	00:22
Diagnostics Check		OK	00:22

Data Download Date/Time: 02/25/2016 00:14

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006214

INSTRUMENT REGISTERED TO: OCALA POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOEDICKE,WILLIAM,R
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: JOEDICKE,WILLIAM
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 01/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:57

SUBJECT'S LAST NAME: JOLLO

DATE/TIME OF TEST: 01/21/2016 00:30:34

Diagnostics Check		OK	00:24
Air Blank		0.000	00:24
Control Test		0.081	00:25
Air Blank		0.000	00:25
Subject Sample	# 1	0.143	00:26
Breath Volume		3.964	00:26
Air Blank		0.000	00:26
Air Blank		0.000	00:28
Subject Sample	# 2	0.146	00:29
Breath Volume		4.027	00:29
Air Blank		0.000	00:29
Control Test		0.080	00:29
Air Blank		0.000	00:30
Diagnostics Check		OK	00:30

Data Download Date/Time: 02/25/2016 00:14

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006214

INSTRUMENT REGISTERED TO: OCALA POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOEDICKE,WILLIAM,R
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: JOEDICKE,WILLIAM
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 01/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:18

SUBJECT'S LAST NAME: GUZMAN

DATE/TIME OF TEST: 01/29/2016 22:59:23

Diagnostics Check		OK	22:49
Air Blank		0.000	22:50
Control Test		0.079	22:50
Air Blank		0.000	22:50
Subject Sample	# 1	0.238	22:51
Breath Volume		1.523	22:51
Air Blank		0.000	22:52
Air Blank		0.000	22:53
Subject Sample	# 2	0.274	22:54
Breath Volume		1.843	22:54
Air Blank		0.000	22:55
Air Blank		0.000	22:56
Subject Sample	# 3	0.264	22:57
Breath Volume		1.867	22:57
Air Blank		0.000	22:58
Control Test		0.078	22:58
Air Blank		0.000	22:59
Diagnostics Check		OK	22:59

Data Download Date/Time: 02/25/2016 00:14

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROZYCKI,RICHARD,J
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: BIVINS,HEATHER
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 01/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG425202
DRY GAS EXPIRY DATE: 09/09/2016
VIOLATION CODE: DUI
ARREST TIME: 02:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:26

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 01/23/2016 02:56:44

Diagnostics Check		OK	02:50
Air Blank		0.000	02:50
Control Test		0.077	02:50
Air Blank		0.000	02:51
Subject Sample	# 1	0.247	02:52
Breath Volume		4.117	02:52
Air Blank		0.000	02:52
Air Blank		0.000	02:54
Subject Sample	# 2	0.255	02:55
Breath Volume		4.269	02:55
Air Blank		0.000	02:55
Control Test		0.075	02:56
Air Blank		0.000	02:56
Diagnostics Check		OK	02:56

Data Download Date/Time: 02/27/2016 18:03

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAGNER,ROBERT,F
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: WAGNER,ROBERT
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 01/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG425202
DRY GAS EXPIRY DATE: 09/09/2016
VIOLATION CODE: COURT ORD
ARREST TIME: 10:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:05

SUBJECT'S LAST NAME: PHILLIPS

DATE/TIME OF TEST: 01/24/2016 10:35:11

Diagnostics Check		OK	10:29	
Air Blank		0.000	10:29	
Control Test		0.077	10:29	
Air Blank		0.000	10:30	
Subject Sample	# 1	0.000	10:30	
Breath Volume		2.761	10:30	
Air Blank		0.000	10:31	
Air Blank		0.000	10:33	
Subject Sample	# 2	0.000	10:33	
Breath Volume		2.480	10:33	
Air Blank		0.000	10:34	
Control Test		0.073	10:34	Control Outside Tolerance
Air Blank		0.000	10:35	

Data Download Date/Time: 02/27/2016 18:03

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARRITY,THOMAS,A
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: GARRITY,THOMAS
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 01/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG425202
DRY GAS EXPIRY DATE: 09/09/2016
VIOLATION CODE: DUI
ARREST TIME: 00:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: YORK

DATE/TIME OF TEST: 01/31/2016 01:42:07

Diagnostics Check		OK	01:38	
Air Blank		0.000	01:38	
Control Test		0.076	01:38	
Air Blank		0.000	01:39	
Subject Sample	# 1		01:40	Subject Test Refused
Breath Volume		0.000	01:40	
Air Blank		0.000	01:41	
Control Test		0.077	01:41	
Air Blank		0.000	01:42	
Diagnostics Check		OK	01:42	

Data Download Date/Time: 02/27/2016 18:03

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006234

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MALDONADO,RYAN,A
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: MALDONADO,RYAN
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1694353
DRY GAS EXPIRY DATE: 06/10/2017
VIOLATION CODE: DUI
ARREST TIME: 00:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: VIDAL-CASIMIRO

DATE/TIME OF TEST: 01/01/2016 01:12:59

Diagnostics Check		OK	01:06
Air Blank		0.000	01:06
Control Test		0.081	01:07
Air Blank		0.000	01:07
Subject Sample	# 1	0.123	01:08
Breath Volume		1.363	01:08
Air Blank		0.000	01:08
Air Blank		0.000	01:10
Subject Sample	# 2	0.136	01:11
Breath Volume		1.187	01:11
Air Blank		0.000	01:12
Control Test		0.080	01:12
Air Blank		0.000	01:12
Diagnostics Check		OK	01:12

Data Download Date/Time: 01/29/2016 13:29

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006234

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REED,S,D
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: SPEAK,V
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1694353
DRY GAS EXPIRY DATE: 06/10/2017
VIOLATION CODE: DUI
ARREST TIME: 15:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:08

SUBJECT'S LAST NAME: PUIG

DATE/TIME OF TEST: 01/01/2016 15:38:23

Diagnostics Check		OK	15:35	
Air Blank		0.000	15:36	
Control Test		0.080	15:36	
Air Blank		0.000	15:36	
Subject Sample	# 1		15:37	Subject Test Refused
Breath Volume		0.000	15:37	
Air Blank		0.000	15:37	
Control Test		0.080	15:37	
Air Blank		0.000	15:38	
Diagnostics Check		OK	15:38	

Data Download Date/Time: 01/29/2016 13:29

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006234

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REED,S,D
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: PASCHER,B
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1694353
DRY GAS EXPIRY DATE: 06/10/2017
VIOLATION CODE: DUI
ARREST TIME: 10:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:00

SUBJECT'S LAST NAME: MIRANDA

DATE/TIME OF TEST: 01/02/2016 11:43:33

Diagnostics Check		OK	11:33
Air Blank		0.000	11:34
Control Test		0.081	11:34
Air Blank		0.000	11:35
Subject Sample	# 1	0.266	11:35
Breath Volume		2.421	11:35
Air Blank		0.000	11:36
Air Blank		0.000	11:38
Subject Sample	# 2	0.235	11:38
Breath Volume		1.109	11:38
Air Blank		0.000	11:39
Air Blank		0.000	11:41
Subject Sample	# 3	0.242	11:41
Breath Volume		1.203	11:41
Air Blank		0.000	11:42
Control Test		0.077	11:42
Air Blank		0.000	11:43
Diagnostics Check		OK	11:43

Data Download Date/Time: 01/29/2016 13:29

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006234

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REED,S,D
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: SPEAK,V
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1694353
DRY GAS EXPIRY DATE: 06/10/2017
VIOLATION CODE: DUI
ARREST TIME: 14:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:20

SUBJECT'S LAST NAME: FLETCHER

DATE/TIME OF TEST: 01/15/2016 14:52:14

Diagnostics Check		OK	14:46	
Air Blank		0.000	14:47	
Control Test		0.081	14:47	
Air Blank		0.000	14:48	
Subject Sample	# 1		14:50	Subject Test Refused
Breath Volume		0.000	14:50	
Air Blank		0.000	14:51	
Control Test		0.081	14:51	
Air Blank		0.000	14:52	
Diagnostics Check		OK	14:52	

Data Download Date/Time: 01/29/2016 13:29

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006234

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAYLOR,MICHAEL,R
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: TAYLOR,ANDY
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1694353
DRY GAS EXPIRY DATE: 06/10/2017
VIOLATION CODE: DUI
ARREST TIME: 03:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:19

SUBJECT'S LAST NAME: TEHUINTLE

DATE/TIME OF TEST: 01/22/2016 04:46:44

Diagnostics Check		OK	04:39
Air Blank		0.000	04:40
Control Test		0.080	04:40
Air Blank		0.000	04:41
Subject Sample	# 1	0.159	04:42
Breath Volume		2.976	04:42
Air Blank		0.000	04:42
Air Blank		0.000	04:44
Subject Sample	# 2	0.158	04:45
Breath Volume		2.996	04:45
Air Blank		0.000	04:45
Control Test		0.077	04:46
Air Blank		0.000	04:46
Diagnostics Check		OK	04:46

Data Download Date/Time: 01/29/2016 13:29

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006234

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MALDONADO,RYAN,A
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: MALDONADO,RYAN
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1694353
DRY GAS EXPIRY DATE: 06/10/2017
VIOLATION CODE: DUI
ARREST TIME: 02:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:44

SUBJECT'S LAST NAME: ARRIETA

DATE/TIME OF TEST: 01/23/2016 03:17:59

Diagnostics Check		OK	03:11
Air Blank		0.000	03:11
Control Test		0.081	03:12
Air Blank		0.000	03:12
Subject Sample	# 1	0.165	03:13
Breath Volume		1.660	03:13
Air Blank		0.000	03:14
Air Blank		0.000	03:15
Subject Sample	# 2	0.174	03:16
Breath Volume		1.652	03:16
Air Blank		0.000	03:17
Control Test		0.082	03:17
Air Blank		0.000	03:17
Diagnostics Check		OK	03:17

Data Download Date/Time: 01/29/2016 13:29

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006234

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MALDONADO,RYAN,A
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: MALDONADO,RYAN
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1694353
DRY GAS EXPIRY DATE: 06/10/2017
VIOLATION CODE: DUI
ARREST TIME: 21:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:54

SUBJECT'S LAST NAME: STUDLEY

DATE/TIME OF TEST: 01/23/2016 22:24:31

Diagnostics Check		OK	22:21	
Air Blank		0.000	22:22	
Control Test		0.082	22:22	
Air Blank		0.000	22:22	
Subject Sample	# 1		22:23	Subject Test Refused
Breath Volume		0.000	22:23	
Air Blank		0.000	22:23	
Control Test		0.082	22:23	
Air Blank		0.000	22:24	
Diagnostics Check		OK	22:24	

Data Download Date/Time: 01/29/2016 13:29

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006234

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MALDONADO,RYAN,A
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: MALDONADO,RYAN
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1694353
DRY GAS EXPIRY DATE: 06/10/2017
VIOLATION CODE: DUI
ARREST TIME: 23:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:12

SUBJECT'S LAST NAME: WHYSALL NIXON

DATE/TIME OF TEST: 01/24/2016 00:43:58

Diagnostics Check		OK	00:41	
Air Blank		0.000	00:41	
Control Test		0.083	00:41	
Air Blank		0.000	00:42	
Subject Sample	# 1		00:42	Subject Test Refused
Breath Volume		0.000	00:42	
Air Blank		0.000	00:43	
Control Test		0.083	00:43	
Air Blank		0.000	00:43	
Diagnostics Check		OK	00:43	

Data Download Date/Time: 01/29/2016 13:29

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006234

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOODS,JOSHUA,H
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: SERRANO,R
ARRESTING OFFICER AGENCY: Seminole Dept of LE
LAST AGENCY INSPECTION: 01/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1694353
DRY GAS EXPIRY DATE: 06/10/2017
VIOLATION CODE: DUI
ARREST TIME: 13:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:00

SUBJECT'S LAST NAME: JIMMIE

DATE/TIME OF TEST: 01/29/2016 15:27:41

Diagnostics Check		OK	15:21
Air Blank		0.000	15:21
Control Test		0.082	15:22
Air Blank		0.000	15:22
Subject Sample	# 1	0.212	15:23
Breath Volume		2.328	15:23
Air Blank		0.000	15:23
Air Blank		0.000	15:25
Subject Sample	# 2	0.208	15:26
Breath Volume		1.515	15:26
Air Blank		0.000	15:26
Control Test		0.081	15:27
Air Blank		0.000	15:27
Diagnostics Check		OK	15:27

Data Download Date/Time: 02/29/2016 14:59

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006234

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPBELL,ADAM,M
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: CAMPBELL,ADAM
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 01/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1694353
DRY GAS EXPIRY DATE: 06/10/2017
VIOLATION CODE: DUI
ARREST TIME: 21:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:18

SUBJECT'S LAST NAME: GAMEZ

DATE/TIME OF TEST: 01/30/2016 23:08:18

Diagnostics Check		OK	23:02
Air Blank		0.000	23:02
Control Test		0.081	23:02
Air Blank		0.000	23:03
Subject Sample	# 1	0.194	23:03
Breath Volume		1.964	23:03
Air Blank		0.000	23:04
Air Blank		0.000	23:06
Subject Sample	# 2	0.183	23:06
Breath Volume		1.777	23:06
Air Blank		0.000	23:07
Control Test		0.080	23:07
Air Blank		0.000	23:08
Diagnostics Check		OK	23:08

Data Download Date/Time: 02/29/2016 14:59

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006234

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPBELL,ADAM,M
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: NORWOOD,CHRISTOPHER
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 01/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1694353
DRY GAS EXPIRY DATE: 06/10/2017
VIOLATION CODE: DUI
ARREST TIME: 21:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:19

SUBJECT'S LAST NAME: VASQUEZ

DATE/TIME OF TEST: 01/30/2016 23:52:49

Diagnostics Check		OK	23:40	
Air Blank		0.000	23:41	
Control Test		0.082	23:41	
Air Blank		0.000	23:41	
Subject Sample	# 1	0.217	23:42	
Breath Volume		1.265	23:42	
Air Blank		0.000	23:43	
Air Blank		0.000	23:44	
Subject Sample	# 2	0.163	23:47	Volume Not Met
Breath Volume		0.488	23:47	
Air Blank		0.000	23:48	
Air Blank		0.000	23:50	
Subject Sample	# 3	0.200	23:51	
Breath Volume		1.402	23:51	
Air Blank		0.000	23:51	
Control Test		0.077	23:52	
Air Blank		0.000	23:52	
Diagnostics Check		OK	23:52	

Data Download Date/Time: 02/29/2016 14:59

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1841081
DRY GAS EXPIRY DATE: 04/08/2018
VIOLATION CODE: DUI
ARREST TIME: 21:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:42

SUBJECT'S LAST NAME: POPPELL JR

DATE/TIME OF TEST: 01/02/2016 23:12:33

Diagnostics Check		OK	23:03	
Air Blank		0.000	23:04	
Control Test		0.079	23:04	
Air Blank		0.000	23:05	
Subject Sample	# 1		23:08	No Sample Provided
Breath Volume		0.000	23:08	
Air Blank		0.000	23:08	
Air Blank		0.000	23:10	
Subject Sample	# 2		23:11	Subject Test Refused
Breath Volume		0.000	23:11	
Air Blank		0.000	23:11	
Control Test		0.079	23:11	
Air Blank		0.000	23:12	
Diagnostics Check		OK	23:12	

Data Download Date/Time: 01/25/2016 09:54

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PETERSON,STEFANIE,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: PETERSON,STEFANIE
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1841081
DRY GAS EXPIRY DATE: 04/08/2018
VIOLATION CODE: DUI
ARREST TIME: 22:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:53

SUBJECT'S LAST NAME: DELGADO-RICO

DATE/TIME OF TEST: 01/04/2016 00:20:39

Diagnostics Check		OK	00:13
Air Blank		0.000	00:14
Control Test		0.080	00:14
Air Blank		0.000	00:15
Subject Sample	# 1	0.184	00:15
Breath Volume		1.250	00:15
Air Blank		0.000	00:16
Air Blank		0.000	00:18
Subject Sample	# 2	0.201	00:18
Breath Volume		2.453	00:18
Air Blank		0.000	00:19
Control Test		0.079	00:20
Air Blank		0.000	00:20
Diagnostics Check		OK	00:20

Data Download Date/Time: 01/25/2016 09:54

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1841081
DRY GAS EXPIRY DATE: 04/08/2018
VIOLATION CODE: DUI
ARREST TIME: 16:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:55

SUBJECT'S LAST NAME: JEWELL JR

DATE/TIME OF TEST: 01/06/2016 18:22:35

Diagnostics Check		OK	18:16
Air Blank		0.000	18:16
Control Test		0.080	18:16
Air Blank		0.000	18:17
Subject Sample	# 1	0.186	18:17
Breath Volume		2.703	18:17
Air Blank		0.000	18:18
Air Blank		0.000	18:20
Subject Sample	# 2	0.189	18:20
Breath Volume		2.679	18:20
Air Blank		0.000	18:21
Control Test		0.075	18:21
Air Blank		0.000	18:22
Diagnostics Check		OK	18:22

Data Download Date/Time: 01/25/2016 09:54

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PETERSON,STEFANIE,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: PETERSON,STEFANIE
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1841081
DRY GAS EXPIRY DATE: 04/08/2018
VIOLATION CODE: DUI
ARREST TIME: 18:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:31

SUBJECT'S LAST NAME: PEXA

DATE/TIME OF TEST: 01/06/2016 20:02:14

Diagnostics Check		OK	19:55
Air Blank		0.000	19:55
Control Test		0.080	19:55
Air Blank		0.000	19:56
Subject Sample	# 1	0.182	19:57
Breath Volume		3.042	19:57
Air Blank		0.000	19:57
Air Blank		0.000	19:59
Subject Sample	# 2	0.170	20:00
Breath Volume		2.359	20:00
Air Blank		0.000	20:01
Control Test		0.079	20:01
Air Blank		0.000	20:02
Diagnostics Check		OK	20:02

Data Download Date/Time: 01/25/2016 09:54

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1841081
DRY GAS EXPIRY DATE: 04/08/2018
VIOLATION CODE: DUI
ARREST TIME: 21:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: VERDIN

DATE/TIME OF TEST: 01/06/2016 22:37:20

Diagnostics Check		OK	22:34	
Air Blank		0.000	22:34	
Control Test		0.079	22:35	
Air Blank		0.000	22:35	
Subject Sample	# 1		22:35	Subject Test Refused
Breath Volume		0.000	22:35	
Air Blank		0.000	22:36	
Control Test		0.079	22:36	
Air Blank		0.000	22:37	
Diagnostics Check		OK	22:37	

Data Download Date/Time: 01/25/2016 09:54

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1841081
DRY GAS EXPIRY DATE: 04/08/2018
VIOLATION CODE: DUI
ARREST TIME: 01:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:13

SUBJECT'S LAST NAME: METZGER

DATE/TIME OF TEST: 01/07/2016 02:36:26

Diagnostics Check		OK	02:33	
Air Blank		0.000	02:34	
Control Test		0.080	02:34	
Air Blank		0.000	02:34	
Subject Sample	# 1		02:35	Subject Test Refused
Breath Volume		0.000	02:35	
Air Blank		0.000	02:35	
Control Test		0.079	02:35	
Air Blank			02:35	RFI Detect
Air Blank		0.000	02:36	

Data Download Date/Time: 01/25/2016 09:54

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COROMINAS,BENNY,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: COROMINAS,BENNY
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1841081
DRY GAS EXPIRY DATE: 04/08/2018
VIOLATION CODE: DUI
ARREST TIME: 11:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: AURISANO

DATE/TIME OF TEST: 01/09/2016 00:27:37

Diagnostics Check		OK	00:20	
Air Blank		0.000	00:20	
Control Test		0.078	00:21	
Air Blank		0.000	00:21	
Subject Sample	# 1	0.142	00:23	
Breath Volume		2.964	00:23	
Air Blank		0.000	00:23	
Air Blank		0.000	00:25	
Subject Sample	# 2	0.134	00:25	
Breath Volume		2.515	00:25	
Air Blank		0.000	00:26	
Control Test		0.073	00:27	Control Outside Tolerance
Air Blank		0.000	00:27	

Data Download Date/Time: 01/25/2016 09:54

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TATE,ROBERT,J
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: JONES,JOHN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1841081
DRY GAS EXPIRY DATE: 04/08/2018
VIOLATION CODE: DUI
ARREST TIME: 02:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 01/09/2016 04:26:47

Diagnostics Check		OK	04:19
Air Blank		0.000	04:20
Control Test		0.079	04:20
Air Blank		0.000	04:21
Subject Sample	# 1	0.192	04:21
Breath Volume		1.433	04:21
Air Blank		0.000	04:22
Air Blank		0.000	04:24
Subject Sample	# 2	0.191	04:25
Breath Volume		1.523	04:25
Air Blank		0.000	04:25
Control Test		0.078	04:26
Air Blank		0.000	04:26
Diagnostics Check		OK	04:26

Data Download Date/Time: 01/25/2016 09:54

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TATE,ROBERT,J
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: TATE,ROBERT
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1841081
DRY GAS EXPIRY DATE: 04/08/2018
VIOLATION CODE: DUI
ARREST TIME: 20:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:10

SUBJECT'S LAST NAME: PRIOR

DATE/TIME OF TEST: 01/09/2016 21:39:00

Diagnostics Check		OK	21:32
Air Blank		0.000	21:32
Control Test		0.079	21:33
Air Blank		0.000	21:33
Subject Sample	# 1	0.000	21:34
Breath Volume		1.906	21:34
Air Blank		0.000	21:34
Air Blank		0.000	21:36
Subject Sample	# 2	0.000	21:37
Breath Volume		2.707	21:37
Air Blank		0.000	21:38
Control Test		0.077	21:38
Air Blank		0.000	21:38
Diagnostics Check		OK	21:39

Data Download Date/Time: 01/25/2016 09:54

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROWE,BRANDON,D
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: ROWE,BRANDON
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1841081
DRY GAS EXPIRY DATE: 04/08/2018
VIOLATION CODE: DUI
ARREST TIME: 23:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:06

SUBJECT'S LAST NAME: SALINAS

DATE/TIME OF TEST: 01/10/2016 00:37:34

Diagnostics Check		OK	00:30
Air Blank		0.000	00:31
Control Test		0.080	00:31
Air Blank		0.000	00:32
Subject Sample	# 1	0.182	00:32
Breath Volume		2.183	00:32
Air Blank		0.000	00:33
Air Blank		0.000	00:35
Subject Sample	# 2	0.181	00:35
Breath Volume		2.179	00:35
Air Blank		0.000	00:36
Control Test		0.078	00:36
Air Blank		0.000	00:37
Diagnostics Check		OK	00:37

Data Download Date/Time: 01/25/2016 09:54

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DE LIMA,ALEX,
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: DE LIMA,ALEX
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1841081
DRY GAS EXPIRY DATE: 04/08/2018
VIOLATION CODE: DUI
ARREST TIME: 13:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:48

SUBJECT'S LAST NAME: MCMURRAY

DATE/TIME OF TEST: 01/10/2016 15:15:08

Diagnostics Check		OK	15:08
Air Blank		0.000	15:09
Control Test		0.079	15:09
Air Blank		0.000	15:09
Subject Sample	# 1	0.202	15:10
Breath Volume		1.890	15:10
Air Blank		0.000	15:11
Air Blank		0.000	15:12
Subject Sample	# 2	0.197	15:13
Breath Volume		2.105	15:13
Air Blank		0.000	15:14
Control Test		0.076	15:14
Air Blank		0.000	15:15
Diagnostics Check		OK	15:15

Data Download Date/Time: 01/25/2016 09:54

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HULL,LINDA,L
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: RAGOONANAN,R
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1841081
DRY GAS EXPIRY DATE: 04/08/2018
VIOLATION CODE: DUI
ARREST TIME: 20:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:16

SUBJECT'S LAST NAME: HULL

DATE/TIME OF TEST: 01/14/2016 22:04:51

Diagnostics Check		OK	21:56
Air Blank		0.000	21:56
Control Test		0.079	21:57
Air Blank		0.000	21:57
Subject Sample	# 1	0.177	21:59
Breath Volume		1.886	21:59
Air Blank		0.000	22:00
Air Blank		0.000	22:01
Subject Sample	# 2	0.175	22:03
Breath Volume		2.671	22:03
Air Blank		0.000	22:03
Control Test		0.077	22:04
Air Blank		0.000	22:04
Diagnostics Check		OK	22:04

Data Download Date/Time: 01/25/2016 09:54

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGOONANAN,R,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: RAGOONANAN,R
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1841081
DRY GAS EXPIRY DATE: 04/08/2018
VIOLATION CODE: DUI
ARREST TIME: 20:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: HULL

DATE/TIME OF TEST: 01/14/2016 22:32:38

Diagnostics Check		OK	22:26
Air Blank		0.000	22:26
Control Test		0.078	22:26
Air Blank		0.000	22:27
Subject Sample	# 1	0.171	22:27
Breath Volume		2.160	22:27
Air Blank		0.000	22:28
Air Blank		0.000	22:30
Subject Sample	# 2	0.174	22:30
Breath Volume		2.585	22:30
Air Blank		0.000	22:31
Control Test		0.079	22:32
Air Blank		0.000	22:32
Diagnostics Check		OK	22:32

Data Download Date/Time: 01/25/2016 09:54

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1841081
DRY GAS EXPIRY DATE: 04/08/2018
VIOLATION CODE: DUI
ARREST TIME: 01:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: BEHORI

DATE/TIME OF TEST: 01/16/2016 03:17:24

Diagnostics Check		OK	03:10
Air Blank		0.000	03:11
Control Test		0.080	03:11
Air Blank		0.000	03:11
Subject Sample	# 1	0.128	03:12
Breath Volume		2.476	03:12
Air Blank		0.000	03:13
Air Blank		0.000	03:15
Subject Sample	# 2	0.124	03:15
Breath Volume		2.406	03:15
Air Blank		0.000	03:16
Control Test		0.075	03:16
Air Blank		0.000	03:17
Diagnostics Check		OK	03:17

Data Download Date/Time: 01/25/2016 09:54

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VONADOR,JERARD,E
BREATH TEST OPERATOR AGENCY: Bradenton PD
ARRESTING OFFICER: VONADOR,JERARD
ARRESTING OFFICER AGENCY: Bradenton PD
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1841081
DRY GAS EXPIRY DATE: 04/08/2018
VIOLATION CODE: DUI
ARREST TIME: 01:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: SCHLOUGH

DATE/TIME OF TEST: 01/16/2016 03:46:44

Diagnostics Check		OK	03:40
Air Blank		0.000	03:40
Control Test		0.079	03:41
Air Blank		0.000	03:41
Subject Sample	# 1	0.160	03:42
Breath Volume		2.667	03:42
Air Blank		0.000	03:42
Air Blank		0.000	03:44
Subject Sample	# 2	0.151	03:45
Breath Volume		2.421	03:45
Air Blank		0.000	03:45
Control Test		0.078	03:46
Air Blank		0.000	03:46
Diagnostics Check		OK	03:46

Data Download Date/Time: 01/25/2016 09:54

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1841081
DRY GAS EXPIRY DATE: 04/08/2018
VIOLATION CODE: DUI
ARREST TIME: 23:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:39

SUBJECT'S LAST NAME: JEFFERIES JR

DATE/TIME OF TEST: 01/17/2016 01:05:42

Diagnostics Check		OK	00:59
Air Blank		0.000	00:59
Control Test		0.079	01:00
Air Blank		0.000	01:00
Subject Sample	# 1	0.206	01:01
Breath Volume		3.785	01:01
Air Blank		0.000	01:01
Air Blank		0.000	01:03
Subject Sample	# 2	0.213	01:04
Breath Volume		3.843	01:04
Air Blank		0.000	01:04
Control Test		0.076	01:05
Air Blank		0.000	01:05
Diagnostics Check		OK	01:05

Data Download Date/Time: 01/25/2016 09:54

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: SANTANA,ELIER
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1841081
DRY GAS EXPIRY DATE: 04/08/2018
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: ARITA

DATE/TIME OF TEST: 01/17/2016 04:31:57

Diagnostics Check		OK	04:26	
Air Blank		0.000	04:26	
Control Test		0.079	04:26	
Air Blank		0.000	04:27	
Subject Sample	# 1	0.245	04:30	Volume Not Met
Breath Volume		1.148	04:30	
Air Blank			04:31	Purge Fail
Air Blank		0.000	04:31	

Data Download Date/Time: 01/25/2016 09:54

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MIKLUSCAK,ZACHARY,J
BREATH TEST OPERATOR AGENCY: Bradenton PD
ARRESTING OFFICER: MIKLUSCAK,ZACHARY
ARRESTING OFFICER AGENCY: Bradenton PD
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1841081
DRY GAS EXPIRY DATE: 04/08/2018
VIOLATION CODE: DUI
ARREST TIME: 23:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:31

SUBJECT'S LAST NAME: ESTEP

DATE/TIME OF TEST: 01/18/2016 01:59:58

Diagnostics Check		OK	01:53
Air Blank		0.000	01:53
Control Test		0.080	01:54
Air Blank		0.000	01:54
Subject Sample	# 1	0.108	01:55
Breath Volume		2.066	01:55
Air Blank		0.000	01:56
Air Blank		0.000	01:57
Subject Sample	# 2	0.103	01:58
Breath Volume		1.992	01:58
Air Blank		0.000	01:59
Control Test		0.079	01:59
Air Blank		0.000	01:59
Diagnostics Check		OK	01:59

Data Download Date/Time: 01/25/2016 09:54

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COROMINAS,BENNY,
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: COROMINAS,BENNY
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1841081
DRY GAS EXPIRY DATE: 04/08/2018
VIOLATION CODE: DUI
ARREST TIME: 21:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:28

SUBJECT'S LAST NAME: LHINEY

DATE/TIME OF TEST: 01/18/2016 22:55:17

Diagnostics Check		OK	22:48
Air Blank		0.000	22:48
Control Test		0.079	22:49
Air Blank		0.000	22:49
Subject Sample	# 1	0.156	22:50
Breath Volume		1.511	22:50
Air Blank		0.000	22:50
Air Blank		0.000	22:52
Subject Sample	# 2	0.159	22:53
Breath Volume		1.285	22:53
Air Blank		0.000	22:54
Control Test		0.075	22:54
Air Blank		0.000	22:55
Diagnostics Check		OK	22:55

Data Download Date/Time: 01/25/2016 09:54

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSON,LEON,J
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: JOHNSON,LEON
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1841081
DRY GAS EXPIRY DATE: 04/08/2018
VIOLATION CODE: DUI
ARREST TIME: 00:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: CLARK

DATE/TIME OF TEST: 01/23/2016 00:41:37

Diagnostics Check	OK	00:40	
Air Blank		00:41	RFI Detect
Air Blank	0.000	00:41	

Data Download Date/Time: 01/25/2016 09:54

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSON,LEON,J
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: JOHNSON,LEON
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1841081
DRY GAS EXPIRY DATE: 04/08/2018
VIOLATION CODE: DUI
ARREST TIME: 00:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: CLARK

DATE/TIME OF TEST: 01/23/2016 00:45:25

Diagnostics Check	OK	00:44	
Air Blank		00:44	RFI Detect
Air Blank	0.000	00:45	

Data Download Date/Time: 01/25/2016 09:54

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSON,LEON,J
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: JOHNSON,LEON
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1841081
DRY GAS EXPIRY DATE: 04/08/2018
VIOLATION CODE: DUI
ARREST TIME: 00:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: CLARK

DATE/TIME OF TEST: 01/23/2016 01:01:21

Diagnostics Check		OK	00:53
Air Blank		0.000	00:54
Control Test		0.080	00:54
Air Blank		0.000	00:55
Subject Sample	# 1	0.186	00:56
Breath Volume		2.097	00:56
Air Blank		0.000	00:57
Air Blank		0.000	00:58
Subject Sample	# 2	0.186	00:59
Breath Volume		2.109	00:59
Air Blank		0.000	01:00
Control Test		0.075	01:00
Air Blank		0.000	01:01
Diagnostics Check		OK	01:01

Data Download Date/Time: 01/25/2016 09:54

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TATE,ROBERT,J
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: TATE,ROBERT
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1841081
DRY GAS EXPIRY DATE: 04/08/2018
VIOLATION CODE: DUI
ARREST TIME: 00:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:58

SUBJECT'S LAST NAME: LAMPHRON

DATE/TIME OF TEST: 01/24/2016 01:31:05

Diagnostics Check		OK	01:22
Air Blank		0.000	01:23
Control Test		0.080	01:23
Air Blank		0.000	01:24
Subject Sample	# 1	0.132	01:26
Breath Volume		1.183	01:26
Air Blank		0.000	01:26
Air Blank		0.000	01:28
Subject Sample	# 2	0.150	01:29
Breath Volume		1.328	01:29
Air Blank		0.000	01:30
Control Test		0.076	01:30
Air Blank		0.000	01:31
Diagnostics Check		OK	01:31

Data Download Date/Time: 01/25/2016 09:54

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ACKERMAN,TYLER,J
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: ACKERMAN,TYLER
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1841081
DRY GAS EXPIRY DATE: 04/08/2018
VIOLATION CODE: DUI
ARREST TIME: 07:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:45

SUBJECT'S LAST NAME: CANIL

DATE/TIME OF TEST: 01/24/2016 11:28:57

Diagnostics Check		OK	11:25	
Air Blank		0.000	11:26	
Control Test		0.080	11:26	
Air Blank		0.000	11:27	
Subject Sample	# 1		11:27	Subject Test Refused
Breath Volume		0.000	11:27	
Air Blank		0.000	11:28	
Control Test		0.080	11:28	
Air Blank		0.000	11:28	
Diagnostics Check		OK	11:28	

Data Download Date/Time: 01/25/2016 09:54

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAKAI,MICHAEL,T
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: MAKAI,MICHAEL
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1841081
DRY GAS EXPIRY DATE: 04/08/2018
VIOLATION CODE: DUI
ARREST TIME: 19:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:05

SUBJECT'S LAST NAME: HERNANDEZ-SIERRA

DATE/TIME OF TEST: 01/24/2016 20:39:41

Diagnostics Check		OK	20:33
Air Blank		0.000	20:33
Control Test		0.080	20:33
Air Blank		0.000	20:34
Subject Sample	# 1	0.224	20:35
Breath Volume		1.265	20:35
Air Blank		0.000	20:35
Air Blank		0.000	20:37
Subject Sample	# 2	0.212	20:38
Breath Volume		1.765	20:38
Air Blank		0.000	20:38
Control Test		0.079	20:39
Air Blank		0.000	20:39
Diagnostics Check		OK	20:39

Data Download Date/Time: 01/25/2016 09:54

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TATE,ROBERT,J
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: TATE,ROBERT
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1841081
DRY GAS EXPIRY DATE: 04/08/2018
VIOLATION CODE: DUI
ARREST TIME: 23:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:12

SUBJECT'S LAST NAME: AUGUSTIN

DATE/TIME OF TEST: 01/25/2016 00:40:09

Diagnostics Check		OK	00:33	
Air Blank		0.000	00:33	
Control Test		0.080	00:34	
Air Blank		0.000	00:34	
Subject Sample	# 1	0.183	00:35	
Breath Volume		2.242	00:35	
Air Blank		0.000	00:36	
Air Blank		0.000	00:38	
Subject Sample	# 2	0.180	00:38	
Breath Volume		2.359	00:38	
Air Blank		0.000	00:39	
Control Test		0.068	00:39	Control Outside Tolerance
Air Blank		0.000	00:40	

Data Download Date/Time: 01/25/2016 09:54

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TATE,ROBERT,J
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: TATE,ROBERT
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1841081
DRY GAS EXPIRY DATE: 04/08/2018
VIOLATION CODE: DUI
ARREST TIME: 23:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:12

SUBJECT'S LAST NAME: AUGUSTIN

DATE/TIME OF TEST: 01/25/2016 00:50:26

Diagnostics Check		OK	00:43
Air Blank		0.000	00:44
Control Test		0.077	00:44
Air Blank		0.000	00:45
Subject Sample	# 1	0.185	00:45
Breath Volume		1.960	00:45
Air Blank		0.000	00:46
Air Blank		0.000	00:48
Subject Sample	# 2	0.185	00:48
Breath Volume		2.273	00:48
Air Blank		0.000	00:49
Control Test		0.078	00:49
Air Blank		0.000	00:50
Diagnostics Check		OK	00:50

Data Download Date/Time: 01/25/2016 09:54

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STACY,A,L
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: MASON,M
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 12/21/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12014080A1
DRY GAS EXPIRY DATE: 06/01/2016
VIOLATION CODE: DUI
ARREST TIME: 00:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: MCRAE

DATE/TIME OF TEST: 01/01/2016 01:10:20

Diagnostics Check		OK	01:03
Air Blank		0.000	01:04
Control Test		0.080	01:04
Air Blank		0.000	01:05
Subject Sample	# 1	0.181	01:05
Breath Volume		2.394	01:05
Air Blank		0.000	01:06
Air Blank		0.000	01:08
Subject Sample	# 2	0.179	01:08
Breath Volume		2.000	01:08
Air Blank		0.000	01:09
Control Test		0.080	01:09
Air Blank		0.000	01:10
Diagnostics Check		OK	01:10

Data Download Date/Time: 01/21/2016 20:31

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZIER,TAMMY,L
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: PRITCHARD,TRAVIS
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 12/21/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12014080A1
DRY GAS EXPIRY DATE: 06/01/2016
VIOLATION CODE: DUI
ARREST TIME: 21:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:29

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 01/02/2016 22:59:28

Diagnostics Check		OK	22:51
Air Blank		0.000	22:52
Control Test		0.080	22:52
Air Blank		0.000	22:53
Subject Sample	# 1	0.173	22:54
Breath Volume		1.332	22:54
Air Blank		0.000	22:55
Air Blank		0.000	22:56
Subject Sample	# 2	0.167	22:57
Breath Volume		1.808	22:57
Air Blank		0.000	22:58
Control Test		0.080	22:58
Air Blank		0.000	22:59
Diagnostics Check		OK	22:59

Data Download Date/Time: 01/21/2016 20:31

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZIER,TAMMY,L
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: CRUZ,C
ARRESTING OFFICER AGENCY: Orange Park PD
LAST AGENCY INSPECTION: 12/21/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12014080A1
DRY GAS EXPIRY DATE: 06/01/2016
VIOLATION CODE: DUI
ARREST TIME: 02:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:45

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 01/07/2016 05:17:08

Diagnostics Check		OK	05:08	
Air Blank		0.000	05:09	
Control Test		0.081	05:09	
Air Blank		0.000	05:09	
Subject Sample	# 1		05:13	No Sample Provided
Breath Volume		0.000	05:13	
Air Blank		0.000	05:13	
Air Blank		0.000	05:15	
Subject Sample	# 2		05:15	Subject Test Refused
Breath Volume		0.000	05:15	
Air Blank		0.000	05:16	
Control Test		0.080	05:16	
Air Blank		0.000	05:17	
Diagnostics Check		OK	05:17	

Data Download Date/Time: 01/21/2016 20:31

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STACY,A,L
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: ROOT,K
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 12/21/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12014080A1
DRY GAS EXPIRY DATE: 06/01/2016
VIOLATION CODE: DUI
ARREST TIME: 22:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: MEDINA

DATE/TIME OF TEST: 01/11/2016 00:46:41

Diagnostics Check		OK	00:38	
Air Blank		0.000	00:38	
Control Test		0.080	00:38	
Air Blank		0.000	00:39	
Subject Sample	# 1		00:42	No Sample Provided
Breath Volume		0.000	00:42	
Air Blank		0.000	00:43	
Air Blank		0.000	00:45	
Subject Sample	# 2		00:45	Subject Test Refused
Breath Volume		0.000	00:45	
Air Blank		0.000	00:45	
Control Test		0.080	00:46	
Air Blank		0.000	00:46	
Diagnostics Check		OK	00:46	

Data Download Date/Time: 01/21/2016 20:31

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZIER,TAMMY,L
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: PRITCHARD,TRTAVIS
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 12/21/2015

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 12014080A1
DRY GAS EXPIRY DATE: 06/01/2016
VIOLATION CODE: DUI
ARREST TIME: 21:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 01/16/2016 22:42:17

Diagnostics Check		OK	22:34
Air Blank		0.000	22:35
Control Test		0.080	22:35
Air Blank		0.000	22:36
Subject Sample	# 1	0.080	22:37
Breath Volume		1.792	22:37
Air Blank		0.000	22:37
Air Blank		0.000	22:39
Subject Sample	# 2	0.080	22:40
Breath Volume		1.554	22:40
Air Blank		0.000	22:41
Control Test		0.080	22:41
Air Blank		0.000	22:42
Diagnostics Check		OK	22:42

Data Download Date/Time: 01/21/2016 20:31

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZIER,TAMMY,L
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: PRITCHARD,TRAVIS
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 12/21/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12014080A1
DRY GAS EXPIRY DATE: 06/01/2016
VIOLATION CODE: DUI
ARREST TIME: 21:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:37

SUBJECT'S LAST NAME: ALVEY

DATE/TIME OF TEST: 01/17/2016 22:10:40

Diagnostics Check		OK	22:04
Air Blank		0.000	22:04
Control Test		0.080	22:05
Air Blank		0.000	22:05
Subject Sample	# 1	0.227	22:06
Breath Volume		3.472	22:06
Air Blank		0.000	22:06
Air Blank		0.000	22:08
Subject Sample	# 2	0.221	22:09
Breath Volume		3.062	22:09
Air Blank		0.000	22:09
Control Test		0.079	22:10
Air Blank		0.000	22:10
Diagnostics Check		OK	22:10

Data Download Date/Time: 01/21/2016 20:31

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZIER,TAMMY,L
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: PRITCHARD,TRAVIS
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 12/21/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12014080A1
DRY GAS EXPIRY DATE: 06/01/2016
VIOLATION CODE: DUI
ARREST TIME: 03:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:19

SUBJECT'S LAST NAME: KOESTENS

DATE/TIME OF TEST: 01/18/2016 04:49:03

Diagnostics Check		OK	04:41
Air Blank		0.000	04:42
Control Test		0.081	04:42
Air Blank		0.000	04:43
Subject Sample	# 1	0.185	04:43
Breath Volume		1.902	04:43
Air Blank		0.000	04:44
Air Blank		0.000	04:46
Subject Sample	# 2	0.168	04:47
Breath Volume		1.453	04:47
Air Blank		0.000	04:48
Control Test		0.081	04:48
Air Blank		0.000	04:48
Diagnostics Check		OK	04:49

Data Download Date/Time: 01/21/2016 20:31

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STACY,A,L
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: SAUNDERS,J
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 12/21/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12014080A1
DRY GAS EXPIRY DATE: 06/01/2016
VIOLATION CODE: DUI
ARREST TIME: 21:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:21

SUBJECT'S LAST NAME: WITTE

DATE/TIME OF TEST: 01/19/2016 22:54:03

Diagnostics Check		OK	22:46
Air Blank		0.000	22:47
Control Test		0.082	22:47
Air Blank		0.000	22:47
Subject Sample	# 1	0.240	22:49
Breath Volume		1.828	22:49
Air Blank		0.000	22:49
Air Blank		0.000	22:51
Subject Sample	# 2	0.242	22:52
Breath Volume		1.843	22:52
Air Blank		0.000	22:53
Control Test		0.081	22:53
Air Blank		0.000	22:53
Diagnostics Check		OK	22:54

Data Download Date/Time: 01/21/2016 20:31

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: DOERR,CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A3
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:56

SUBJECT'S LAST NAME: JIMENEZ

DATE/TIME OF TEST: 01/01/2016 00:25:27

Diagnostics Check		OK	00:19
Air Blank		0.000	00:19
Control Test		0.081	00:20
Air Blank		0.000	00:20
Subject Sample	# 1	0.089	00:21
Breath Volume		2.425	00:21
Air Blank		0.000	00:21
Air Blank		0.000	00:23
Subject Sample	# 2	0.082	00:23
Breath Volume		3.632	00:23
Air Blank		0.000	00:24
Control Test		0.078	00:24
Air Blank		0.000	00:25
Diagnostics Check		OK	00:25

Data Download Date/Time: 01/11/2016 13:49

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MAYO,RICKY
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A3
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:44

SUBJECT'S LAST NAME: BASSAN

DATE/TIME OF TEST: 01/01/2016 03:13:50

Diagnostics Check		OK	03:07
Air Blank		0.000	03:07
Control Test		0.081	03:07
Air Blank		0.000	03:08
Subject Sample	# 1	0.094	03:09
Breath Volume		2.062	03:09
Air Blank		0.000	03:09
Air Blank		0.000	03:11
Subject Sample	# 2	0.092	03:12
Breath Volume		1.988	03:12
Air Blank		0.000	03:12
Control Test		0.080	03:13
Air Blank		0.000	03:13
Diagnostics Check		OK	03:13

Data Download Date/Time: 01/11/2016 13:49

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: GONZALEZ CORTINA,RAUL
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A3
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:55

SUBJECT'S LAST NAME: SHORT HAMIWKA

DATE/TIME OF TEST: 01/01/2016 05:33:32

Diagnostics Check		OK	05:20	
Air Blank		0.000	05:20	
Control Test		0.080	05:21	
Air Blank		0.000	05:21	
Subject Sample	# 1	0.189	05:24	Volume Not Met
Breath Volume		0.015	05:24	
Air Blank		0.000	05:25	
Air Blank		0.000	05:27	
Subject Sample	# 2	0.189	05:28	
Breath Volume		3.039	05:28	
Air Blank		0.000	05:29	
Air Blank		0.000	05:30	
Subject Sample	# 3	0.178	05:31	
Breath Volume		2.359	05:31	
Air Blank		0.000	05:32	
Control Test		0.078	05:32	
Air Blank		0.000	05:33	
Diagnostics Check		OK	05:33	

Data Download Date/Time: 01/11/2016 13:49

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: GRUBER,MICHAEL
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A3
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 18:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:01

SUBJECT'S LAST NAME: MORALES

DATE/TIME OF TEST: 01/01/2016 19:51:56

Diagnostics Check		OK	19:45
Air Blank		0.000	19:45
Control Test		0.080	19:45
Air Blank		0.000	19:46
Subject Sample	# 1	0.219	19:47
Breath Volume		1.437	19:47
Air Blank		0.000	19:48
Air Blank		0.000	19:49
Subject Sample	# 2	0.221	19:50
Breath Volume		1.480	19:50
Air Blank		0.000	19:51
Control Test		0.080	19:51
Air Blank		0.000	19:51
Diagnostics Check		OK	19:51

Data Download Date/Time: 01/11/2016 13:49

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MCCOY,JOSHUA
ARRESTING OFFICER AGENCY: North Palm Beach
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A3
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 18:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:54

SUBJECT'S LAST NAME: WINDLE

DATE/TIME OF TEST: 01/02/2016 19:26:44

Diagnostics Check		OK	19:20
Air Blank		0.000	19:20
Control Test		0.081	19:20
Air Blank		0.000	19:21
Subject Sample	# 1	0.105	19:22
Breath Volume		1.574	19:22
Air Blank		0.000	19:22
Air Blank		0.000	19:24
Subject Sample	# 2	0.100	19:25
Breath Volume		1.640	19:25
Air Blank		0.000	19:25
Control Test		0.080	19:26
Air Blank		0.000	19:26
Diagnostics Check		OK	19:26

Data Download Date/Time: 01/11/2016 13:49

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: KWAKU,EDWARD
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A3
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:43

SUBJECT'S LAST NAME: BRUCELLA

DATE/TIME OF TEST: 01/03/2016 01:17:28

Diagnostics Check		OK	01:11
Air Blank		0.000	01:11
Control Test		0.081	01:11
Air Blank		0.000	01:12
Subject Sample	# 1	0.000	01:13
Breath Volume		2.664	01:13
Air Blank		0.000	01:13
Air Blank		0.000	01:15
Subject Sample	# 2	0.000	01:16
Breath Volume		2.968	01:16
Air Blank		0.000	01:16
Control Test		0.080	01:16
Air Blank		0.000	01:17
Diagnostics Check		OK	01:17

Data Download Date/Time: 01/11/2016 13:49

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHAFER,NANCY,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BORROWS,ANDREW
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A3
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:56

SUBJECT'S LAST NAME: NORMANDALE

DATE/TIME OF TEST: 01/03/2016 21:36:57

Diagnostics Check		OK	21:22	
Air Blank		0.000	21:22	
Control Test		0.080	21:22	
Air Blank		0.000	21:23	
Subject Sample	# 1	0.236	21:26	Volume Not Met
Breath Volume		0.914	21:26	
Air Blank		0.000	21:27	
Air Blank		0.000	21:29	
Subject Sample	# 2	0.248	21:29	
Breath Volume		1.386	21:29	
Air Blank		0.000	21:30	
Air Blank		0.000	21:32	
Subject Sample	# 3	0.222	21:35	Slope Not Level
Breath Volume		1.265	21:35	
Air Blank		0.000	21:35	
Control Test		0.079	21:36	
Air Blank		0.000	21:36	
Diagnostics Check		OK	21:36	

Data Download Date/Time: 01/11/2016 13:49

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHAFER,NANCY,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BORROWS,ANDREW
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A3
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:56

SUBJECT'S LAST NAME: NORMANDALE

DATE/TIME OF TEST: 01/03/2016 21:51:09

Diagnostics Check		OK	21:39	
Air Blank		0.000	21:40	
Control Test		0.077	21:40	
Air Blank		0.000	21:41	
Subject Sample	# 1	0.200	21:44	Volume Not Met
Breath Volume		0.765	21:44	
Air Blank		0.000	21:44	
Air Blank		0.000	21:46	
Subject Sample	# 2		21:49	No Sample Provided
Breath Volume		0.000	21:49	
Air Blank		0.000	21:50	
Control Test		0.077	21:50	
Air Blank		0.000	21:51	
Diagnostics Check		OK	21:51	

Data Download Date/Time: 01/11/2016 13:49

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CHAHINE,DEREK
ARRESTING OFFICER AGENCY: Delray Beach PD
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A3
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:26

SUBJECT'S LAST NAME: HABOUTO

DATE/TIME OF TEST: 01/06/2016 04:02:21

Diagnostics Check		OK	03:52
Air Blank		0.000	03:53
Control Test		0.081	03:53
Air Blank		0.000	03:54
Subject Sample	# 1	0.166	03:56
Breath Volume		1.324	03:56
Air Blank		0.000	03:56
Air Blank		0.000	03:58
Subject Sample	# 2	0.173	04:00
Breath Volume		1.519	04:00
Air Blank		0.000	04:01
Control Test		0.079	04:01
Air Blank		0.000	04:02
Diagnostics Check		OK	04:02

Data Download Date/Time: 01/11/2016 13:49

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: DOERR,CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A3
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:07

SUBJECT'S LAST NAME: MASSEY

DATE/TIME OF TEST: 01/07/2016 00:36:30

Diagnostics Check		OK	00:30
Air Blank		0.000	00:30
Control Test		0.080	00:31
Air Blank		0.000	00:31
Subject Sample	# 1	0.160	00:32
Breath Volume		2.339	00:32
Air Blank		0.000	00:32
Air Blank		0.000	00:34
Subject Sample	# 2	0.157	00:34
Breath Volume		2.359	00:34
Air Blank		0.000	00:35
Control Test		0.080	00:35
Air Blank		0.000	00:36
Diagnostics Check		OK	00:36

Data Download Date/Time: 01/11/2016 13:49

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAIN,JEANETTE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: GONZALEZ-CORTINA,R H
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A3
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: SCHWARTZ

DATE/TIME OF TEST: 01/08/2016 00:22:50

Diagnostics Check		OK	00:16
Air Blank		0.000	00:16
Control Test		0.080	00:16
Air Blank		0.000	00:17
Subject Sample	# 1	0.123	00:18
Breath Volume		1.203	00:18
Air Blank		0.000	00:18
Air Blank		0.000	00:20
Subject Sample	# 2	0.120	00:21
Breath Volume		1.667	00:21
Air Blank		0.000	00:21
Control Test		0.081	00:22
Air Blank		0.000	00:22
Diagnostics Check		OK	00:22

Data Download Date/Time: 01/11/2016 13:49

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MAYO,RICKY
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A3
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: ROBERTS

DATE/TIME OF TEST: 01/08/2016 01:43:28

Diagnostics Check		OK	01:35
Air Blank		0.000	01:35
Control Test		0.080	01:36
Air Blank		0.000	01:36
Subject Sample	# 1	0.215	01:38
Breath Volume		2.144	01:38
Air Blank		0.000	01:38
Air Blank		0.000	01:40
Subject Sample	# 2	0.204	01:41
Breath Volume		1.832	01:41
Air Blank		0.000	01:42
Control Test		0.078	01:42
Air Blank		0.000	01:43
Diagnostics Check		OK	01:43

Data Download Date/Time: 01/11/2016 13:49

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SOLOWAY,ALAN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A3
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:21

SUBJECT'S LAST NAME: BOSS

DATE/TIME OF TEST: 01/08/2016 03:55:52

Diagnostics Check		OK	03:47
Air Blank		0.000	03:47
Control Test		0.080	03:47
Air Blank		0.000	03:48
Subject Sample	# 1	0.217	03:49
Breath Volume		4.804	03:49
Air Blank		0.000	03:50
Air Blank		0.000	03:52
Subject Sample	# 2	0.213	03:54
Breath Volume		4.421	03:54
Air Blank		0.000	03:54
Control Test		0.079	03:55
Air Blank		0.000	03:55
Diagnostics Check		OK	03:55

Data Download Date/Time: 01/11/2016 13:49

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FICARRA,CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A3
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:11

SUBJECT'S LAST NAME: LATRONICA

DATE/TIME OF TEST: 01/09/2016 03:42:12

Diagnostics Check		OK	03:35
Air Blank		0.000	03:36
Control Test		0.081	03:36
Air Blank		0.000	03:37
Subject Sample	# 1	0.122	03:37
Breath Volume		2.101	03:37
Air Blank		0.000	03:38
Air Blank		0.000	03:40
Subject Sample	# 2	0.120	03:40
Breath Volume		2.000	03:40
Air Blank		0.000	03:41
Control Test		0.080	03:41
Air Blank		0.000	03:42
Diagnostics Check		OK	03:42

Data Download Date/Time: 01/11/2016 13:49

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WARD,CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A3
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:04

SUBJECT'S LAST NAME: YEPEZ

DATE/TIME OF TEST: 01/09/2016 05:33:44

Diagnostics Check		OK	05:27
Air Blank		0.000	05:27
Control Test		0.080	05:28
Air Blank		0.000	05:28
Subject Sample	# 1	0.145	05:29
Breath Volume		1.769	05:29
Air Blank		0.000	05:29
Air Blank		0.000	05:31
Subject Sample	# 2	0.145	05:32
Breath Volume		1.644	05:32
Air Blank		0.000	05:32
Control Test		0.080	05:33
Air Blank		0.000	05:33
Diagnostics Check		OK	05:33

Data Download Date/Time: 01/11/2016 13:49

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAIN,JEANETTE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SENTMANAT,A
ARRESTING OFFICER AGENCY: Greenacres Pub Sf
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A3
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 19:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:50

SUBJECT'S LAST NAME: ARTEAGA

DATE/TIME OF TEST: 01/09/2016 20:21:16

Diagnostics Check		OK	20:14
Air Blank		0.000	20:14
Control Test		0.080	20:14
Air Blank		0.000	20:15
Subject Sample	# 1	0.340	20:16
Breath Volume		1.207	20:16
Air Blank		0.000	20:17
Air Blank		0.000	20:19
Subject Sample	# 2	0.339	20:19
Breath Volume		1.660	20:19
Air Blank		0.000	20:20
Control Test		0.080	20:20
Air Blank		0.000	20:21
Diagnostics Check		OK	20:21

Data Download Date/Time: 01/11/2016 13:49

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARI,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LASSO,JAVIER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A3
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:02

SUBJECT'S LAST NAME: ABREGO

DATE/TIME OF TEST: 01/10/2016 04:38:09

Diagnostics Check		OK	04:31
Air Blank		0.000	04:31
Control Test		0.080	04:31
Air Blank		0.000	04:32
Subject Sample	# 1	0.241	04:33
Breath Volume		1.128	04:33
Air Blank		0.000	04:33
Air Blank		0.000	04:35
Subject Sample	# 2	0.233	04:36
Breath Volume		1.113	04:36
Air Blank		0.000	04:37
Control Test		0.079	04:37
Air Blank		0.000	04:38
Diagnostics Check		OK	04:38

Data Download Date/Time: 01/11/2016 13:49

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHAFER,NANCY,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SCHNEIDER,JOSHUA
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A3
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 10:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:03

SUBJECT'S LAST NAME: MCCARTHY

DATE/TIME OF TEST: 01/11/2016 11:34:55

Diagnostics Check		OK	11:27
Air Blank		0.000	11:27
Control Test		0.081	11:28
Air Blank		0.000	11:28
Subject Sample	# 1	0.329	11:29
Breath Volume		2.042	11:29
Air Blank		0.000	11:30
Air Blank		0.000	11:32
Subject Sample	# 2	0.320	11:33
Breath Volume		2.023	11:33
Air Blank		0.000	11:33
Control Test		0.079	11:34
Air Blank		0.000	11:34
Diagnostics Check		OK	11:34

Data Download Date/Time: 01/11/2016 13:49

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ADDEA,ROBERT
ARRESTING OFFICER AGENCY: Delray Beach PD
LAST AGENCY INSPECTION: 01/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1870563
DRY GAS EXPIRY DATE: 06/17/2018
VIOLATION CODE: DUI
ARREST TIME: 04:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:55

SUBJECT'S LAST NAME: FEIG

DATE/TIME OF TEST: 01/12/2016 05:23:56

Diagnostics Check		OK	05:17
Air Blank		0.000	05:18
Control Test		0.081	05:18
Air Blank		0.000	05:18
Subject Sample	# 1	0.037	05:19
Breath Volume		4.562	05:19
Air Blank		0.000	05:20
Air Blank		0.000	05:21
Subject Sample	# 2	0.036	05:22
Breath Volume		4.558	05:22
Air Blank		0.000	05:23
Control Test		0.081	05:23
Air Blank		0.000	05:23
Diagnostics Check		OK	05:23

Data Download Date/Time: 02/17/2016 12:02

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DORSEY,PAMELA,F
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CHIQUITO-RODRIGUEZ,LENNDY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 01/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1870563
DRY GAS EXPIRY DATE: 06/17/2018
VIOLATION CODE: DUI
ARREST TIME: 22:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: MARTE-PEREZ

DATE/TIME OF TEST: 01/12/2016 23:34:54

Diagnostics Check		OK	23:27
Air Blank		0.000	23:28
Control Test		0.081	23:28
Air Blank		0.000	23:29
Subject Sample	# 1	0.209	23:30
Breath Volume		1.968	23:30
Air Blank		0.000	23:30
Air Blank		0.000	23:32
Subject Sample	# 2	0.206	23:33
Breath Volume		2.210	23:33
Air Blank		0.000	23:33
Control Test		0.080	23:34
Air Blank		0.000	23:34
Diagnostics Check		OK	23:34

Data Download Date/Time: 02/17/2016 12:02

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LEBLANC,ARIELLE
ARRESTING OFFICER AGENCY: Delray Beach PD
LAST AGENCY INSPECTION: 01/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1870563
DRY GAS EXPIRY DATE: 06/17/2018
VIOLATION CODE: DUI
ARREST TIME: 00:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:56

SUBJECT'S LAST NAME: CHAHALIS

DATE/TIME OF TEST: 01/13/2016 02:40:29

Diagnostics Check		OK	02:33
Air Blank		0.000	02:34
Control Test		0.081	02:34
Air Blank		0.000	02:35
Subject Sample	# 1	0.134	02:35
Breath Volume		2.871	02:35
Air Blank		0.000	02:36
Air Blank		0.000	02:38
Subject Sample	# 2	0.131	02:38
Breath Volume		2.558	02:38
Air Blank		0.000	02:39
Control Test		0.080	02:39
Air Blank		0.000	02:40
Diagnostics Check		OK	02:40

Data Download Date/Time: 02/17/2016 12:02

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DORSEY,PAMELA,F
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BORROWS,ANDREW
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 01/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1870563
DRY GAS EXPIRY DATE: 06/17/2018
VIOLATION CODE: DUI
ARREST TIME: 19:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:30

SUBJECT'S LAST NAME: KOZLIN

DATE/TIME OF TEST: 01/13/2016 21:05:25

Diagnostics Check		OK	20:58
Air Blank		0.000	20:58
Control Test		0.081	20:58
Air Blank		0.000	20:59
Subject Sample	# 1	0.149	21:00
Breath Volume		1.394	21:00
Air Blank		0.000	21:01
Air Blank		0.000	21:03
Subject Sample	# 2	0.155	21:03
Breath Volume		1.476	21:03
Air Blank		0.000	21:04
Control Test		0.081	21:04
Air Blank		0.000	21:05
Diagnostics Check		OK	21:05

Data Download Date/Time: 02/17/2016 12:02

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DORSEY,PAMELA,F
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: GUBISH,MATTHEW
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 01/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1870563
DRY GAS EXPIRY DATE: 06/17/2018
VIOLATION CODE: DUI
ARREST TIME: 20:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:03

SUBJECT'S LAST NAME: SANDS

DATE/TIME OF TEST: 01/13/2016 21:43:30

Diagnostics Check		OK	21:36
Air Blank		0.000	21:37
Control Test		0.080	21:37
Air Blank		0.000	21:38
Subject Sample	# 1	0.179	21:38
Breath Volume		2.296	21:38
Air Blank		0.000	21:39
Air Blank		0.000	21:41
Subject Sample	# 2	0.182	21:41
Breath Volume		1.296	21:41
Air Blank		0.000	21:42
Control Test		0.081	21:42
Air Blank		0.000	21:43
Diagnostics Check		OK	21:43

Data Download Date/Time: 02/17/2016 12:02

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DORSEY,PAMELA,F
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: GONZALEZ-CORTINA,RAUL
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 01/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1870563
DRY GAS EXPIRY DATE: 06/17/2018
VIOLATION CODE: DUI
ARREST TIME: 21:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:09

SUBJECT'S LAST NAME: RAMIREZ-MORALES

DATE/TIME OF TEST: 01/13/2016 22:44:51

Diagnostics Check		OK	22:36
Air Blank		0.000	22:37
Control Test		0.081	22:37
Air Blank		0.000	22:38
Subject Sample	# 1	0.284	22:39
Breath Volume		2.191	22:39
Air Blank		0.000	22:40
Air Blank		0.000	22:42
Subject Sample	# 2	0.276	22:43
Breath Volume		2.074	22:43
Air Blank		0.000	22:43
Control Test		0.080	22:44
Air Blank		0.000	22:44
Diagnostics Check		OK	22:44

Data Download Date/Time: 02/17/2016 12:02

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FICARRA,CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 01/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1870563
DRY GAS EXPIRY DATE: 06/17/2018
VIOLATION CODE: DUI
ARREST TIME: 03:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: LOPEZ ACEYTUNO

DATE/TIME OF TEST: 01/17/2016 04:27:23

Diagnostics Check		OK	04:18
Air Blank		0.000	04:19
Control Test		0.080	04:19
Air Blank		0.000	04:20
Subject Sample	# 1	0.195	04:22
Breath Volume		2.367	04:22
Air Blank		0.000	04:23
Air Blank		0.000	04:25
Subject Sample	# 2	0.193	04:25
Breath Volume		2.250	04:25
Air Blank		0.000	04:26
Control Test		0.078	04:26
Air Blank		0.000	04:27
Diagnostics Check		OK	04:27

Data Download Date/Time: 02/17/2016 12:02

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MERCIER,DANIEL
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 01/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1870563
DRY GAS EXPIRY DATE: 06/17/2018
VIOLATION CODE: DUI
ARREST TIME: 04:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:33

SUBJECT'S LAST NAME: WALKER

DATE/TIME OF TEST: 01/17/2016 05:07:26

Diagnostics Check		OK	05:01
Air Blank		0.000	05:01
Control Test		0.080	05:02
Air Blank		0.000	05:02
Subject Sample	# 1	0.164	05:02
Breath Volume		2.214	05:02
Air Blank		0.000	05:03
Air Blank		0.000	05:05
Subject Sample	# 2	0.157	05:05
Breath Volume		2.031	05:05
Air Blank		0.000	05:06
Control Test		0.080	05:06
Air Blank		0.000	05:07
Diagnostics Check		OK	05:07

Data Download Date/Time: 02/17/2016 12:02

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: HEGELE,BRANDON
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 01/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1870563
DRY GAS EXPIRY DATE: 06/17/2018
VIOLATION CODE: DUI
ARREST TIME: 06:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:17

SUBJECT'S LAST NAME: YANZANIS

DATE/TIME OF TEST: 01/19/2016 07:47:57

Diagnostics Check		OK	07:40
Air Blank		0.000	07:40
Control Test		0.082	07:41
Air Blank		0.000	07:41
Subject Sample	# 1	0.184	07:42
Breath Volume		1.828	07:42
Air Blank		0.000	07:43
Air Blank		0.000	07:45
Subject Sample	# 2	0.183	07:46
Breath Volume		1.703	07:46
Air Blank		0.000	07:47
Control Test		0.081	07:47
Air Blank		0.000	07:47
Diagnostics Check		OK	07:47

Data Download Date/Time: 02/17/2016 12:02

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LASSO,JAVIER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 01/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1870563
DRY GAS EXPIRY DATE: 06/17/2018
VIOLATION CODE: DUI
ARREST TIME: 02:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:17

SUBJECT'S LAST NAME: GASPAR

DATE/TIME OF TEST: 01/20/2016 03:48:39

Diagnostics Check		OK	03:41
Air Blank		0.000	03:42
Control Test		0.080	03:42
Air Blank		0.000	03:43
Subject Sample	# 1	0.201	03:44
Breath Volume		2.027	03:44
Air Blank		0.000	03:44
Air Blank		0.000	03:46
Subject Sample	# 2	0.195	03:47
Breath Volume		2.917	03:47
Air Blank		0.000	03:47
Control Test		0.081	03:48
Air Blank		0.000	03:48
Diagnostics Check		OK	03:48

Data Download Date/Time: 02/17/2016 12:02

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FICARRA,CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 01/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1870563
DRY GAS EXPIRY DATE: 06/17/2018
VIOLATION CODE: DUI
ARREST TIME: 03:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:31

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 01/21/2016 05:01:13

Diagnostics Check		OK	04:54
Air Blank		0.000	04:55
Control Test		0.082	04:55
Air Blank		0.000	04:55
Subject Sample	# 1	0.149	04:56
Breath Volume		1.640	04:56
Air Blank		0.000	04:57
Air Blank		0.000	04:59
Subject Sample	# 2	0.140	04:59
Breath Volume		1.761	04:59
Air Blank		0.000	05:00
Control Test		0.081	05:00
Air Blank		0.000	05:01
Diagnostics Check		OK	05:01

Data Download Date/Time: 02/17/2016 12:02

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: HILL,JEREMY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 01/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1870563
DRY GAS EXPIRY DATE: 06/17/2018
VIOLATION CODE: DUI
ARREST TIME: 01:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: COEN

DATE/TIME OF TEST: 01/22/2016 02:04:21

Diagnostics Check		OK	01:55
Air Blank		0.000	01:55
Control Test		0.081	01:56
Air Blank		0.000	01:56
Subject Sample	# 1	0.185	01:57
Breath Volume		1.226	01:57
Air Blank		0.000	01:58
Air Blank		0.000	01:59
Subject Sample	# 2	0.186	02:02
Breath Volume		1.324	02:02
Air Blank		0.000	02:03
Control Test		0.081	02:03
Air Blank		0.000	02:04
Diagnostics Check		OK	02:04

Data Download Date/Time: 02/17/2016 12:02

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: DEROGATIS,JOSEPH
ARRESTING OFFICER AGENCY: Palm Springs PD
LAST AGENCY INSPECTION: 01/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1870563
DRY GAS EXPIRY DATE: 06/17/2018
VIOLATION CODE: DUI
ARREST TIME: 03:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: KOC

DATE/TIME OF TEST: 01/23/2016 04:00:37

Diagnostics Check		OK	03:54
Air Blank		0.000	03:54
Control Test		0.081	03:55
Air Blank		0.000	03:55
Subject Sample	# 1	0.149	03:56
Breath Volume		4.015	03:56
Air Blank		0.000	03:56
Air Blank		0.000	03:58
Subject Sample	# 2	0.156	03:59
Breath Volume		3.945	03:59
Air Blank		0.000	03:59
Control Test		0.082	04:00
Air Blank		0.000	04:00
Diagnostics Check		OK	04:00

Data Download Date/Time: 02/17/2016 12:02

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: GONZALEZ-CORTINA,RAUL
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 01/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1870563
DRY GAS EXPIRY DATE: 06/17/2018
VIOLATION CODE: DUI
ARREST TIME: 21:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: COLE

DATE/TIME OF TEST: 01/23/2016 22:21:41

Diagnostics Check		OK	22:15
Air Blank		0.000	22:15
Control Test		0.082	22:15
Air Blank		0.000	22:16
Subject Sample	# 1	0.100	22:16
Breath Volume		2.050	22:16
Air Blank		0.000	22:17
Air Blank		0.000	22:19
Subject Sample	# 2	0.107	22:20
Breath Volume		3.292	22:20
Air Blank		0.000	22:20
Control Test		0.081	22:21
Air Blank		0.000	22:21
Diagnostics Check		OK	22:21

Data Download Date/Time: 02/17/2016 12:02

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: GONZALEZ-CORTINA,RAUL
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 01/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1870563
DRY GAS EXPIRY DATE: 06/17/2018
VIOLATION CODE: DUI
ARREST TIME: 00:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: LAYSON

DATE/TIME OF TEST: 01/24/2016 01:24:34

Diagnostics Check		OK	01:17
Air Blank		0.000	01:17
Control Test		0.082	01:17
Air Blank		0.000	01:18
Subject Sample	# 1	0.000	01:20
Breath Volume		1.246	01:20
Air Blank		0.000	01:20
Air Blank		0.000	01:22
Subject Sample	# 2	0.000	01:23
Breath Volume		1.464	01:23
Air Blank		0.000	01:23
Control Test		0.081	01:23
Air Blank		0.000	01:24
Diagnostics Check		OK	01:24

Data Download Date/Time: 02/17/2016 12:02

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CUMMINGS,PETER
ARRESTING OFFICER AGENCY: Lantana PD
LAST AGENCY INSPECTION: 01/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1870563
DRY GAS EXPIRY DATE: 06/17/2018
VIOLATION CODE: DUI
ARREST TIME: 01:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: MARADIAGA BENITEZ

DATE/TIME OF TEST: 01/24/2016 02:10:06

Diagnostics Check		OK	02:03
Air Blank		0.000	02:03
Control Test		0.081	02:04
Air Blank		0.000	02:04
Subject Sample	# 1	0.124	02:05
Breath Volume		1.222	02:05
Air Blank		0.000	02:06
Air Blank		0.000	02:07
Subject Sample	# 2	0.118	02:08
Breath Volume		1.714	02:08
Air Blank		0.000	02:09
Control Test		0.080	02:09
Air Blank		0.000	02:10
Diagnostics Check		OK	02:10

Data Download Date/Time: 02/17/2016 12:02

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CHIQUITO RODRIGUEZ,LENNDY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 01/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1870563
DRY GAS EXPIRY DATE: 06/17/2018
VIOLATION CODE: DUI
ARREST TIME: 21:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:06

SUBJECT'S LAST NAME: HILL

DATE/TIME OF TEST: 01/24/2016 22:35:10

Diagnostics Check		OK	22:28
Air Blank		0.000	22:29
Control Test		0.082	22:29
Air Blank		0.000	22:29
Subject Sample	# 1	0.173	22:30
Breath Volume		3.843	22:30
Air Blank		0.000	22:31
Air Blank		0.000	22:33
Subject Sample	# 2	0.169	22:33
Breath Volume		3.566	22:33
Air Blank		0.000	22:34
Control Test		0.082	22:34
Air Blank		0.000	22:35
Diagnostics Check		OK	22:35

Data Download Date/Time: 02/17/2016 12:02

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARI,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SCHAEFER,JOHN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 01/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1870563
DRY GAS EXPIRY DATE: 06/17/2018
VIOLATION CODE: DUI
ARREST TIME: 03:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:12

SUBJECT'S LAST NAME: MARQUEZ

DATE/TIME OF TEST: 01/25/2016 04:44:39

Diagnostics Check		OK	04:38
Air Blank		0.000	04:38
Control Test		0.082	04:39
Air Blank		0.000	04:39
Subject Sample	# 1	0.134	04:40
Breath Volume		1.414	04:40
Air Blank		0.000	04:40
Air Blank		0.000	04:42
Subject Sample	# 2	0.141	04:43
Breath Volume		2.144	04:43
Air Blank		0.000	04:43
Control Test		0.081	04:44
Air Blank		0.000	04:44
Diagnostics Check		OK	04:44

Data Download Date/Time: 02/17/2016 12:02

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LEMBO,BRYAN
ARRESTING OFFICER AGENCY: Greenacres Pub Sf
LAST AGENCY INSPECTION: 01/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1870563
DRY GAS EXPIRY DATE: 06/17/2018
VIOLATION CODE: DUI
ARREST TIME: 23:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: JUAN LORENZO

DATE/TIME OF TEST: 01/27/2016 00:45:06

Diagnostics Check		OK	00:31	
Air Blank		0.000	00:32	
Control Test		0.081	00:32	
Air Blank		0.000	00:33	
Subject Sample	# 1	0.252	00:34	
Breath Volume		1.218	00:34	
Air Blank		0.000	00:35	
Air Blank		0.000	00:37	
Subject Sample	# 2	0.219	00:40	Volume Not Met
Breath Volume		0.035	00:40	
Air Blank		0.000	00:41	
Air Blank		0.000	00:42	
Subject Sample	# 3	0.235	00:43	
Breath Volume		1.203	00:43	
Air Blank		0.000	00:44	
Control Test		0.077	00:44	
Air Blank		0.000	00:45	
Diagnostics Check		OK	00:45	

Data Download Date/Time: 02/17/2016 12:02

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: GONZALEZ-CORTINA,RAUL
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 01/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1870563
DRY GAS EXPIRY DATE: 06/17/2018
VIOLATION CODE: DUI
ARREST TIME: 00:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:56

SUBJECT'S LAST NAME: MORRISON

DATE/TIME OF TEST: 01/29/2016 01:24:48

Diagnostics Check		OK	01:19	
Air Blank		0.000	01:19	
Control Test		0.081	01:20	
Air Blank		0.000	01:20	
Subject Sample	# 1	0.240	01:21	
Breath Volume		1.304	01:21	
Air Blank		0.000	01:21	
Air Blank			01:23	RFI Detect
Air Blank		0.000	01:23	
Control Test		0.081	01:24	
Air Blank		0.000	01:24	
Diagnostics Check		OK	01:24	

Data Download Date/Time: 02/17/2016 12:02

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: GONZALEZ-CORTINA,RAUL
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 01/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1870563
DRY GAS EXPIRY DATE: 06/17/2018
VIOLATION CODE: DUI
ARREST TIME: 00:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:56

SUBJECT'S LAST NAME: MORRISON

DATE/TIME OF TEST: 01/29/2016 01:36:11

Diagnostics Check		OK	01:26
Air Blank		0.000	01:26
Control Test		0.081	01:26
Air Blank		0.000	01:27
Subject Sample	# 1	0.245	01:27
Breath Volume		2.007	01:27
Air Blank		0.000	01:28
Air Blank		0.000	01:30
Subject Sample	# 2	0.220	01:31
Breath Volume		1.390	01:31
Air Blank		0.000	01:31
Air Blank		0.000	01:33
Subject Sample	# 3	0.243	01:34
Breath Volume		3.574	01:34
Air Blank		0.000	01:35
Control Test		0.079	01:35
Air Blank		0.000	01:36
Diagnostics Check		OK	01:36

Data Download Date/Time: 02/17/2016 12:02

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHAFER,NANCY,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ESCARAN,L
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 01/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1870563
DRY GAS EXPIRY DATE: 06/17/2018
VIOLATION CODE: DUI
ARREST TIME: 11:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:46

SUBJECT'S LAST NAME: PERRETTE

DATE/TIME OF TEST: 01/29/2016 12:22:15

Diagnostics Check		OK	12:15
Air Blank		0.000	12:16
Control Test		0.081	12:16
Air Blank		0.000	12:17
Subject Sample	# 1	0.000	12:17
Breath Volume		1.980	12:17
Air Blank		0.000	12:18
Air Blank		0.000	12:20
Subject Sample	# 2	0.000	12:20
Breath Volume		1.417	12:20
Air Blank		0.000	12:21
Control Test		0.081	12:21
Air Blank		0.000	12:22
Diagnostics Check		OK	12:22

Data Download Date/Time: 02/17/2016 12:02

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DORSEY,PAMELA,F
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: HANTON,MELINDA
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 01/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1870563
DRY GAS EXPIRY DATE: 06/17/2018
VIOLATION CODE: DUI
ARREST TIME: 21:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:04

SUBJECT'S LAST NAME: HUGHES

DATE/TIME OF TEST: 01/30/2016 22:36:58

Diagnostics Check		OK	22:30
Air Blank		0.000	22:30
Control Test		0.081	22:31
Air Blank		0.000	22:31
Subject Sample	# 1	0.000	22:32
Breath Volume		2.531	22:32
Air Blank		0.000	22:33
Air Blank		0.000	22:35
Subject Sample	# 2	0.000	22:35
Breath Volume		2.230	22:35
Air Blank		0.000	22:36
Control Test		0.081	22:36
Air Blank		0.000	22:36
Diagnostics Check		OK	22:36

Data Download Date/Time: 02/17/2016 12:02

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CHIQUITO-RODRIGUEZ,LENNDY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 01/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1870563
DRY GAS EXPIRY DATE: 06/17/2018
VIOLATION CODE: DUI
ARREST TIME: 02:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: VASQUEZ-LOPEZ

DATE/TIME OF TEST: 01/31/2016 03:59:50

Diagnostics Check		OK	03:48
Air Blank		0.000	03:48
Control Test		0.081	03:49
Air Blank		0.000	03:49
Subject Sample	# 1	0.159	03:51
Breath Volume		2.847	03:51
Air Blank		0.000	03:52
Air Blank		0.000	03:53
Subject Sample	# 2	0.137	03:54
Breath Volume		1.660	03:54
Air Blank		0.000	03:55
Air Blank		0.000	03:57
Subject Sample	# 3	0.143	03:58
Breath Volume		1.718	03:58
Air Blank		0.000	03:58
Control Test		0.080	03:59
Air Blank		0.000	03:59
Diagnostics Check		OK	03:59

Data Download Date/Time: 02/17/2016 12:02

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DORSEY,PAMELA,F
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: TINEO,NELSON
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 01/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1870563
DRY GAS EXPIRY DATE: 06/17/2018
VIOLATION CODE: DUI
ARREST TIME: 19:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: ROSSITTO

DATE/TIME OF TEST: 01/31/2016 21:29:09

Diagnostics Check		OK	21:22
Air Blank		0.000	21:22
Control Test		0.081	21:22
Air Blank		0.000	21:23
Subject Sample	# 1	0.253	21:24
Breath Volume		4.058	21:24
Air Blank		0.000	21:25
Air Blank		0.000	21:26
Subject Sample	# 2	0.262	21:27
Breath Volume		4.812	21:27
Air Blank		0.000	21:28
Control Test		0.080	21:28
Air Blank		0.000	21:29
Diagnostics Check		OK	21:29

Data Download Date/Time: 02/17/2016 12:02

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DORSEY,PAMELA,F
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MERCIER,DANIEL
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203140804A
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: ONOFRE-REYES

DATE/TIME OF TEST: 01/01/2016 02:11:06

Diagnostics Check		OK	02:03
Air Blank		0.000	02:04
Control Test		0.081	02:04
Air Blank		0.000	02:04
Subject Sample	# 1	0.181	02:06
Breath Volume		1.546	02:06
Air Blank		0.000	02:07
Air Blank		0.000	02:08
Subject Sample	# 2	0.191	02:09
Breath Volume		1.609	02:09
Air Blank		0.000	02:10
Control Test		0.081	02:10
Air Blank		0.000	02:11
Diagnostics Check		OK	02:11

Data Download Date/Time: 01/11/2016 12:06

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: STEPHAN,ROBERT
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203140804A
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 02:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:33

SUBJECT'S LAST NAME: FRANCIOSA

DATE/TIME OF TEST: 01/01/2016 03:01:54

Diagnostics Check		OK	02:55
Air Blank		0.000	02:56
Control Test		0.081	02:56
Air Blank		0.000	02:57
Subject Sample	# 1	0.178	02:57
Breath Volume		1.832	02:57
Air Blank		0.000	02:58
Air Blank		0.000	02:59
Subject Sample	# 2	0.189	03:00
Breath Volume		3.253	03:00
Air Blank		0.000	03:00
Control Test		0.081	03:01
Air Blank		0.000	03:01
Diagnostics Check		OK	03:01

Data Download Date/Time: 01/11/2016 12:06

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WALTON,THOMAS
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203140804A
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 04:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:28

SUBJECT'S LAST NAME: MCKAIN

DATE/TIME OF TEST: 01/01/2016 04:57:57

Diagnostics Check		OK	04:51
Air Blank		0.000	04:52
Control Test		0.081	04:52
Air Blank		0.000	04:53
Subject Sample	# 1	0.078	04:53
Breath Volume		1.441	04:53
Air Blank		0.000	04:54
Air Blank		0.000	04:56
Subject Sample	# 2	0.081	04:56
Breath Volume		1.578	04:56
Air Blank		0.000	04:57
Control Test		0.081	04:57
Air Blank		0.000	04:57
Diagnostics Check		OK	04:57

Data Download Date/Time: 01/11/2016 12:06

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CHIQUITO-RODRIGUEZ,LENNDY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203140804A
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 18:26
GENDER: Male

OBSERVATION PERIOD BEGAN: 19:13

SUBJECT'S LAST NAME: GOMEZ-ANDRES

DATE/TIME OF TEST: 01/01/2016 20:31:40

Diagnostics Check		OK	20:24
Air Blank		0.000	20:25
Control Test		0.081	20:25
Air Blank		0.000	20:26
Subject Sample	# 1	0.166	20:26
Breath Volume		1.371	20:26
Air Blank		0.000	20:27
Air Blank		0.000	20:29
Subject Sample	# 2	0.165	20:30
Breath Volume		1.875	20:30
Air Blank		0.000	20:30
Control Test		0.081	20:31
Air Blank		0.000	20:31
Diagnostics Check		OK	20:31

Data Download Date/Time: 01/11/2016 12:06

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: GONZALEZ,ROY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203140804A
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: BAIRD

DATE/TIME OF TEST: 01/03/2016 03:28:12

Diagnostics Check		OK	03:14	
Air Blank		0.000	03:14	
Control Test		0.081	03:14	
Air Blank		0.000	03:15	
Subject Sample	# 1		03:18	No Sample Provided
Breath Volume		0.000	03:18	
Air Blank		0.000	03:19	
Air Blank		0.000	03:21	
Subject Sample	# 2	0.240	03:22	
Breath Volume		1.519	03:22	
Air Blank		0.000	03:22	
Air Blank		0.000	03:24	
Subject Sample	# 3	0.231	03:26	
Breath Volume		2.007	03:26	
Air Blank		0.000	03:27	
Control Test		0.081	03:27	
Air Blank		0.000	03:28	
Diagnostics Check		OK	03:28	

Data Download Date/Time: 01/11/2016 12:06

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SCHNEIDER,JOSHUA
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203140804A
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 06:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:40

SUBJECT'S LAST NAME: ALEXANDRE

DATE/TIME OF TEST: 01/03/2016 07:08:13

Diagnostics Check		OK	07:01
Air Blank		0.000	07:02
Control Test		0.081	07:02
Air Blank		0.000	07:03
Subject Sample	# 1	0.089	07:03
Breath Volume		4.125	07:03
Air Blank		0.000	07:04
Air Blank		0.000	07:06
Subject Sample	# 2	0.090	07:06
Breath Volume		5.117	07:06
Air Blank		0.000	07:07
Control Test		0.081	07:07
Air Blank		0.000	07:08
Diagnostics Check		OK	07:08

Data Download Date/Time: 01/11/2016 12:06

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CASELLA,KATHERINE
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203140804A
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 11:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:14

SUBJECT'S LAST NAME: FLETCHER

DATE/TIME OF TEST: 01/04/2016 14:46:35

Diagnostics Check		OK	14:40
Air Blank		0.000	14:40
Control Test		0.081	14:41
Air Blank		0.000	14:41
Subject Sample	# 1	0.000	14:42
Breath Volume		3.476	14:42
Air Blank		0.000	14:42
Air Blank		0.000	14:44
Subject Sample	# 2	0.000	14:45
Breath Volume		3.699	14:45
Air Blank		0.000	14:45
Control Test		0.082	14:45
Air Blank		0.000	14:46
Diagnostics Check		OK	14:46

Data Download Date/Time: 01/11/2016 12:06

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CROWLEY,PHIL
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203140804A
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 14:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:08

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 01/04/2016 15:36:27

Diagnostics Check		OK	15:30
Air Blank		0.000	15:30
Control Test		0.081	15:31
Air Blank		0.000	15:31
Subject Sample	# 1	0.153	15:31
Breath Volume		1.140	15:31
Air Blank		0.000	15:32
Air Blank		0.000	15:34
Subject Sample	# 2	0.151	15:34
Breath Volume		1.769	15:34
Air Blank		0.000	15:35
Control Test		0.081	15:35
Air Blank		0.000	15:36
Diagnostics Check		OK	15:36

Data Download Date/Time: 01/11/2016 12:06

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DORSEY,PAMELA,F
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: AGUERO,CARLOS
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203140804A
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 17:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:02

SUBJECT'S LAST NAME: OLMOS

DATE/TIME OF TEST: 01/05/2016 18:34:54

Diagnostics Check		OK	18:27
Air Blank		0.000	18:28
Control Test		0.082	18:28
Air Blank		0.000	18:29
Subject Sample	# 1	0.000	18:30
Breath Volume		2.949	18:30
Air Blank		0.000	18:30
Air Blank		0.000	18:32
Subject Sample	# 2	0.000	18:33
Breath Volume		1.800	18:33
Air Blank		0.000	18:33
Control Test		0.081	18:34
Air Blank		0.000	18:34
Diagnostics Check		OK	18:34

Data Download Date/Time: 01/11/2016 12:06

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DORSEY,PAMELA,F
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FREY,JACOB
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203140804A
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 20:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:35

SUBJECT'S LAST NAME: LAFORTE

DATE/TIME OF TEST: 01/05/2016 22:16:37

Diagnostics Check		OK	22:05
Air Blank		0.000	22:06
Control Test		0.081	22:06
Air Blank		0.000	22:06
Subject Sample	# 1	0.157	22:09
Breath Volume		1.332	22:09
Air Blank		0.000	22:10
Air Blank		0.000	22:12
Subject Sample	# 2	0.156	22:14
Breath Volume		1.324	22:14
Air Blank		0.000	22:15
Control Test		0.080	22:16
Air Blank		0.000	22:16
Diagnostics Check		OK	22:16

Data Download Date/Time: 01/11/2016 12:06

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARI,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BEESLEY,CHRISTOPHER
ARRESTING OFFICER AGENCY: Lantana PD
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203140804A
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: YOUNG

DATE/TIME OF TEST: 01/06/2016 01:50:08

Diagnostics Check		OK	01:36	
Air Blank		0.000	01:37	
Control Test		0.081	01:37	
Air Blank		0.000	01:38	
Subject Sample	# 1	0.174	01:41	Volume Not Met
Breath Volume		0.179	01:41	
Air Blank		0.000	01:42	
Air Blank		0.000	01:43	
Subject Sample	# 2	0.173	01:44	
Breath Volume		1.136	01:44	
Air Blank		0.000	01:45	
Air Blank		0.000	01:47	
Subject Sample	# 3	0.164	01:48	
Breath Volume		1.488	01:48	
Air Blank		0.000	01:49	
Control Test		0.079	01:49	
Air Blank		0.000	01:50	
Diagnostics Check		OK	01:50	

Data Download Date/Time: 01/11/2016 12:06

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: HANTON,MELINDA
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203140804A
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 01:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 01/08/2016 02:21:42

Diagnostics Check		OK	02:13
Air Blank		0.000	02:14
Control Test		0.081	02:14
Air Blank		0.000	02:15
Subject Sample	# 1	0.244	02:16
Breath Volume		2.171	02:16
Air Blank		0.000	02:17
Air Blank		0.000	02:18
Subject Sample	# 2	0.239	02:20
Breath Volume		2.398	02:20
Air Blank		0.000	02:20
Control Test		0.080	02:21
Air Blank		0.000	02:21
Diagnostics Check		OK	02:21

Data Download Date/Time: 01/11/2016 12:06

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: DIEZ,DARRELL
ARRESTING OFFICER AGENCY: Palm Springs PD
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203140804A
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 01:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 01/09/2016 02:08:21

Diagnostics Check		OK	02:02
Air Blank		0.000	02:02
Control Test		0.081	02:02
Air Blank		0.000	02:03
Subject Sample	# 1	0.166	02:03
Breath Volume		1.992	02:03
Air Blank		0.000	02:04
Air Blank		0.000	02:06
Subject Sample	# 2	0.165	02:06
Breath Volume		2.093	02:06
Air Blank		0.000	02:07
Control Test		0.081	02:07
Air Blank		0.000	02:08
Diagnostics Check		OK	02:08

Data Download Date/Time: 01/11/2016 12:06

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: DOERR,CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203140804A
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 04:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:35

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 01/09/2016 05:05:19

Diagnostics Check		OK	04:58
Air Blank		0.000	04:59
Control Test		0.082	04:59
Air Blank		0.000	05:00
Subject Sample	# 1	0.049	05:00
Breath Volume		2.832	05:00
Air Blank		0.000	05:01
Air Blank		0.000	05:03
Subject Sample	# 2	0.048	05:03
Breath Volume		2.289	05:03
Air Blank		0.000	05:04
Control Test		0.081	05:04
Air Blank		0.000	05:05
Diagnostics Check		OK	05:05

Data Download Date/Time: 01/11/2016 12:06

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: DIEZ,DARRELL
ARRESTING OFFICER AGENCY: Palm Springs PD
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203140804A
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 02:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: SANDERS

DATE/TIME OF TEST: 01/10/2016 03:39:45

Diagnostics Check		OK	03:32
Air Blank		0.000	03:32
Control Test		0.082	03:33
Air Blank		0.000	03:33
Subject Sample	# 1	0.174	03:34
Breath Volume		2.167	03:34
Air Blank		0.000	03:35
Air Blank		0.000	03:37
Subject Sample	# 2	0.188	03:38
Breath Volume		2.132	03:38
Air Blank		0.000	03:38
Control Test		0.080	03:39
Air Blank		0.000	03:39
Diagnostics Check		OK	03:39

Data Download Date/Time: 01/11/2016 12:06

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: TOUSSAINT,LESLIE
ARRESTING OFFICER AGENCY: Palm Springs PD
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203140804A
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 04:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:14

SUBJECT'S LAST NAME: ORTIZ ALMEIDA

DATE/TIME OF TEST: 01/10/2016 04:51:04

Diagnostics Check		OK	04:38	
Air Blank		0.000	04:38	
Control Test		0.081	04:38	
Air Blank		0.000	04:39	
Subject Sample	# 1	0.238	04:40	
Breath Volume		1.300	04:40	
Air Blank		0.000	04:41	
Air Blank		0.000	04:43	
Subject Sample	# 2	0.216	04:46	Volume Not Met
Breath Volume		0.179	04:46	
Air Blank		0.000	04:47	
Air Blank		0.000	04:48	
Subject Sample	# 3		04:49	Subject Test Refused
Breath Volume		0.000	04:49	
Air Blank		0.000	04:50	
Control Test		0.077	04:50	
Air Blank		0.000	04:50	
Diagnostics Check		OK	04:51	

Data Download Date/Time: 01/11/2016 12:06

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DORSEY,PAMELA,F
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WARD,CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203140804A
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 20:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:40

SUBJECT'S LAST NAME: WOSCHINK

DATE/TIME OF TEST: 01/10/2016 21:12:52

Diagnostics Check		OK	21:05
Air Blank		0.000	21:06
Control Test		0.081	21:06
Air Blank		0.000	21:07
Subject Sample	# 1	0.111	21:08
Breath Volume		3.792	21:08
Air Blank		0.000	21:08
Air Blank		0.000	21:10
Subject Sample	# 2	0.107	21:11
Breath Volume		3.972	21:11
Air Blank		0.000	21:11
Control Test		0.080	21:12
Air Blank		0.000	21:12
Diagnostics Check		OK	21:12

Data Download Date/Time: 01/11/2016 12:06

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FICARRA,CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203140804A
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 01:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:38

SUBJECT'S LAST NAME: GIAMBRONE

DATE/TIME OF TEST: 01/11/2016 02:09:22

Diagnostics Check		OK	02:00
Air Blank		0.000	02:01
Control Test		0.081	02:01
Air Blank		0.000	02:01
Subject Sample	# 1	0.127	02:02
Breath Volume		1.304	02:02
Air Blank		0.000	02:02
Air Blank		0.000	02:04
Subject Sample	# 2	0.110	02:07
Breath Volume		2.105	02:07
Air Blank		0.000	02:08
Control Test		0.081	02:08
Air Blank		0.000	02:09
Diagnostics Check		OK	02:09

Data Download Date/Time: 01/11/2016 12:06

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DORSEY,PAMELA,F
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FROMMER,JOSEPH
ARRESTING OFFICER AGENCY: Lantana PD
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203140804A
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 03:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: MENDEZ VICENTE

DATE/TIME OF TEST: 01/11/2016 03:52:09

Diagnostics Check		OK	03:45
Air Blank		0.000	03:45
Control Test		0.081	03:46
Air Blank		0.000	03:46
Subject Sample	# 1	0.198	03:47
Breath Volume		3.218	03:47
Air Blank		0.000	03:48
Air Blank		0.000	03:49
Subject Sample	# 2	0.198	03:50
Breath Volume		3.835	03:50
Air Blank		0.000	03:51
Control Test		0.080	03:51
Air Blank		0.000	03:52
Diagnostics Check		OK	03:52

Data Download Date/Time: 01/11/2016 12:06

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: TOUSSAINT,LESLIE
ARRESTING OFFICER AGENCY: Palm Springs PD
LAST AGENCY INSPECTION: 01/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203140804A
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 20:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:13

SUBJECT'S LAST NAME: FLORES HERNANDEZ

DATE/TIME OF TEST: 01/11/2016 21:55:34

Diagnostics Check		OK	21:49
Air Blank		0.000	21:49
Control Test		0.081	21:49
Air Blank		0.000	21:50
Subject Sample	# 1	0.311	21:51
Breath Volume		1.382	21:51
Air Blank		0.000	21:51
Air Blank		0.000	21:53
Subject Sample	# 2	0.308	21:53
Breath Volume		1.144	21:53
Air Blank		0.000	21:54
Control Test		0.081	21:54
Air Blank		0.000	21:55
Diagnostics Check		OK	21:55

Data Download Date/Time: 02/17/2016 14:53

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FROMMER,JOSEPH
ARRESTING OFFICER AGENCY: Lantana PD
LAST AGENCY INSPECTION: 01/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1870563
DRY GAS EXPIRY DATE: 06/17/2018
VIOLATION CODE: DUI
ARREST TIME: 21:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: RODRIGUEZ GONZALEZ

DATE/TIME OF TEST: 01/13/2016 22:19:01

Diagnostics Check		OK	22:11
Air Blank		0.000	22:12
Control Test		0.082	22:12
Air Blank		0.000	22:12
Subject Sample	# 1	0.163	22:13
Breath Volume		1.476	22:13
Air Blank		0.000	22:14
Air Blank		0.000	22:15
Subject Sample	# 2	0.165	22:17
Breath Volume		1.933	22:17
Air Blank		0.000	22:18
Control Test		0.082	22:18
Air Blank		0.000	22:18
Diagnostics Check		OK	22:19

Data Download Date/Time: 02/17/2016 14:53

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CHIQUITO-RODRIGUEZ,LENNDY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 01/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1870563
DRY GAS EXPIRY DATE: 06/17/2018
VIOLATION CODE: DUI
ARREST TIME: 23:54
GENDER: Male

OBSERVATION PERIOD BEGAN: 00:37

SUBJECT'S LAST NAME: HOOKS

DATE/TIME OF TEST: 01/15/2016 01:10:43

Diagnostics Check		OK	01:02
Air Blank		0.000	01:03
Control Test		0.081	01:03
Air Blank		0.000	01:03
Subject Sample	# 1	0.086	01:05
Breath Volume		1.253	01:05
Air Blank		0.000	01:06
Air Blank		0.000	01:08
Subject Sample	# 2	0.092	01:09
Breath Volume		1.945	01:09
Air Blank		0.000	01:09
Control Test		0.080	01:10
Air Blank		0.000	01:10
Diagnostics Check		OK	01:10

Data Download Date/Time: 02/17/2016 14:53

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: GONZALEZ-CORTINA,RAUL
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 01/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1870563
DRY GAS EXPIRY DATE: 06/17/2018
VIOLATION CODE: DUI
ARREST TIME: 01:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:36

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 01/15/2016 02:18:44

Diagnostics Check		OK	02:11
Air Blank		0.000	02:11
Control Test		0.081	02:12
Air Blank		0.000	02:12
Subject Sample	# 1	0.163	02:13
Breath Volume		2.261	02:13
Air Blank		0.000	02:14
Air Blank		0.000	02:15
Subject Sample	# 2	0.169	02:17
Breath Volume		3.191	02:17
Air Blank		0.000	02:17
Control Test		0.081	02:18
Air Blank		0.000	02:18
Diagnostics Check		OK	02:18

Data Download Date/Time: 02/17/2016 14:53

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: HANTON,MELINDA
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 01/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1870563
DRY GAS EXPIRY DATE: 06/17/2018
VIOLATION CODE: DUI
ARREST TIME: 21:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:43

SUBJECT'S LAST NAME: FENTON

DATE/TIME OF TEST: 01/16/2016 00:15:35

Diagnostics Check		OK	00:08
Air Blank		0.000	00:08
Control Test		0.081	00:08
Air Blank		0.000	00:09
Subject Sample	# 1	0.176	00:09
Breath Volume		1.117	00:09
Air Blank		0.000	00:10
Air Blank		0.000	00:12
Subject Sample	# 2	0.169	00:13
Breath Volume		1.414	00:13
Air Blank		0.000	00:14
Control Test		0.082	00:14
Air Blank		0.000	00:15
Diagnostics Check		OK	00:15

Data Download Date/Time: 02/17/2016 14:53

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAIN,JEANETTE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FICARRA,CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 01/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1870563
DRY GAS EXPIRY DATE: 06/17/2018
VIOLATION CODE: DUI
ARREST TIME: 21:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:25

SUBJECT'S LAST NAME: ANDRES

DATE/TIME OF TEST: 01/16/2016 22:05:17

Diagnostics Check		OK	21:57
Air Blank		0.000	21:58
Control Test		0.082	21:58
Air Blank		0.000	21:59
Subject Sample	# 1	0.194	22:00
Breath Volume		2.648	22:00
Air Blank		0.000	22:01
Air Blank		0.000	22:03
Subject Sample	# 2	0.202	22:03
Breath Volume		2.824	22:03
Air Blank		0.000	22:04
Control Test		0.080	22:04
Air Blank		0.000	22:05
Diagnostics Check		OK	22:05

Data Download Date/Time: 02/17/2016 14:53

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: GONZALEZ-CORTINA,RAUL
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 01/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1870563
DRY GAS EXPIRY DATE: 06/17/2018
VIOLATION CODE: DUI
ARREST TIME: 02:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:38

SUBJECT'S LAST NAME: MENDEZ

DATE/TIME OF TEST: 01/17/2016 04:17:25

Diagnostics Check		OK	04:07
Air Blank		0.000	04:07
Control Test		0.082	04:07
Air Blank		0.000	04:08
Subject Sample	# 1	0.155	04:11
Breath Volume		1.832	04:11
Air Blank		0.000	04:12
Air Blank		0.000	04:13
Subject Sample	# 2	0.139	04:15
Breath Volume		1.617	04:15
Air Blank		0.000	04:16
Control Test		0.079	04:16
Air Blank		0.000	04:17
Diagnostics Check		OK	04:17

Data Download Date/Time: 02/17/2016 14:53

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: DEROGATIS,JOSEPH
ARRESTING OFFICER AGENCY: Palm Springs PD
LAST AGENCY INSPECTION: 01/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1870563
DRY GAS EXPIRY DATE: 06/17/2018
VIOLATION CODE: DUI
ARREST TIME: 03:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: MENDEZ CARBAJAL

DATE/TIME OF TEST: 01/17/2016 05:04:06

Diagnostics Check		OK	04:55
Air Blank		0.000	04:55
Control Test		0.080	04:55
Air Blank		0.000	04:56
Subject Sample	# 1	0.218	04:59
Breath Volume		2.617	04:59
Air Blank		0.000	04:59
Air Blank		0.000	05:01
Subject Sample	# 2	0.201	05:02
Breath Volume		1.570	05:02
Air Blank		0.000	05:03
Control Test		0.079	05:03
Air Blank		0.000	05:04
Diagnostics Check		OK	05:04

Data Download Date/Time: 02/17/2016 14:53

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SALVEMINI,MICHAEL
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 01/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1870563
DRY GAS EXPIRY DATE: 06/17/2018
VIOLATION CODE: DUI
ARREST TIME: 04:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:21

SUBJECT'S LAST NAME: MENDEZ CAMPOSECO

DATE/TIME OF TEST: 01/17/2016 06:06:49

Diagnostics Check		OK	05:59
Air Blank		0.000	06:00
Control Test		0.081	06:00
Air Blank		0.000	06:01
Subject Sample	# 1	0.228	06:02
Breath Volume		2.476	06:02
Air Blank		0.000	06:02
Air Blank		0.000	06:04
Subject Sample	# 2	0.229	06:05
Breath Volume		2.156	06:05
Air Blank		0.000	06:05
Control Test		0.079	06:06
Air Blank		0.000	06:06
Diagnostics Check		OK	06:06

Data Download Date/Time: 02/17/2016 14:53

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SMITH,MARK
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 01/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1870563
DRY GAS EXPIRY DATE: 06/17/2018
VIOLATION CODE: DUI
ARREST TIME: 01:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: MARTIN JR

DATE/TIME OF TEST: 01/20/2016 02:31:27

Diagnostics Check		OK	02:24
Air Blank		0.000	02:25
Control Test		0.082	02:25
Air Blank		0.000	02:26
Subject Sample	# 1	0.109	02:26
Breath Volume		2.250	02:26
Air Blank		0.000	02:27
Air Blank		0.000	02:29
Subject Sample	# 2	0.111	02:29
Breath Volume		2.785	02:29
Air Blank		0.000	02:30
Control Test		0.082	02:30
Air Blank		0.000	02:31
Diagnostics Check		OK	02:31

Data Download Date/Time: 02/17/2016 14:53

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHAFER,NANCY,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: GODDEN,JOHN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 01/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1870563
DRY GAS EXPIRY DATE: 06/17/2018
VIOLATION CODE: DUI
ARREST TIME: 11:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:45

SUBJECT'S LAST NAME: ALBANIS

DATE/TIME OF TEST: 01/21/2016 13:37:47

Diagnostics Check		OK	13:31
Air Blank		0.000	13:31
Control Test		0.083	13:31
Air Blank		0.000	13:32
Subject Sample	# 1	0.000	13:33
Breath Volume		3.703	13:33
Air Blank		0.000	13:33
Air Blank		0.000	13:35
Subject Sample	# 2	0.000	13:36
Breath Volume		3.179	13:36
Air Blank		0.000	13:36
Control Test		0.082	13:37
Air Blank		0.000	13:37
Diagnostics Check		OK	13:37

Data Download Date/Time: 02/17/2016 14:53

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARI,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ZEITZ,PATRICK
ARRESTING OFFICER AGENCY: Greenacres Pub Sf
LAST AGENCY INSPECTION: 01/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1870563
DRY GAS EXPIRY DATE: 06/17/2018
VIOLATION CODE: DUI
ARREST TIME: 21:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:01

SUBJECT'S LAST NAME: ESCOBAR

DATE/TIME OF TEST: 01/22/2016 22:56:50

Diagnostics Check		OK	22:45	
Air Blank		0.000	22:45	
Control Test		0.082	22:45	
Air Blank		0.000	22:46	
Subject Sample	# 1	0.147	22:49	Volume Not Met
Breath Volume		0.699	22:49	
Air Blank		0.000	22:50	
Air Blank		0.000	22:52	
Subject Sample	# 2	0.129	22:55	Volume Not Met
Breath Volume		0.421	22:55	
Air Blank		0.000	22:55	
Control Test		0.080	22:56	
Air Blank		0.000	22:56	
Diagnostics Check		OK	22:56	

Data Download Date/Time: 02/17/2016 14:53

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARI,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ZEITZ,PATRICK
ARRESTING OFFICER AGENCY: Greenacres Pub Sf
LAST AGENCY INSPECTION: 01/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1870563
DRY GAS EXPIRY DATE: 06/17/2018
VIOLATION CODE: DUI
ARREST TIME: 21:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:01

SUBJECT'S LAST NAME: ESCOBAR

DATE/TIME OF TEST: 01/22/2016 23:10:46

Diagnostics Check		OK	22:58	
Air Blank		0.000	22:59	
Control Test		0.081	22:59	
Air Blank		0.000	23:00	
Subject Sample	# 1	0.142	23:00	
Breath Volume		1.312	23:00	
Air Blank		0.000	23:01	
Air Blank		0.000	23:02	
Subject Sample	# 2	0.128	23:06	Volume Not Met
Breath Volume		0.187	23:06	
Air Blank		0.000	23:06	
Air Blank		0.000	23:08	
Subject Sample	# 3	0.127	23:09	
Breath Volume		1.246	23:09	
Air Blank		0.000	23:09	
Control Test		0.079	23:10	
Air Blank		0.000	23:10	
Diagnostics Check		OK	23:10	

Data Download Date/Time: 02/17/2016 14:53

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAIN,JEANETTE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: TINEO,NELSON
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 01/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1870563
DRY GAS EXPIRY DATE: 06/17/2018
VIOLATION CODE: DUI
ARREST TIME: 18:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:55

SUBJECT'S LAST NAME: COLON

DATE/TIME OF TEST: 01/23/2016 19:27:14

Diagnostics Check		OK	19:19
Air Blank		0.000	19:20
Control Test		0.082	19:20
Air Blank		0.000	19:20
Subject Sample	# 1	0.214	19:22
Breath Volume		2.015	19:22
Air Blank		0.000	19:23
Air Blank		0.000	19:25
Subject Sample	# 2	0.204	19:25
Breath Volume		1.867	19:25
Air Blank		0.000	19:26
Control Test		0.082	19:26
Air Blank		0.000	19:27
Diagnostics Check		OK	19:27

Data Download Date/Time: 02/17/2016 14:53

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAIN,JEANETTE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MALLEY,CHARLES
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 01/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1870563
DRY GAS EXPIRY DATE: 06/17/2018
VIOLATION CODE: DUI
ARREST TIME: 21:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:22

SUBJECT'S LAST NAME: DIGGS

DATE/TIME OF TEST: 01/23/2016 22:00:46

Diagnostics Check		OK	21:45	
Air Blank		0.000	21:45	
Control Test		0.082	21:46	
Air Blank		0.000	21:46	
Subject Sample	# 1	0.166	21:49	Volume Not Met
Breath Volume		0.265	21:49	
Air Blank		0.000	21:50	
Air Blank		0.000	21:52	
Subject Sample	# 2	0.192	21:53	
Breath Volume		1.304	21:53	
Air Blank		0.000	21:54	
Air Blank		0.000	21:56	
Subject Sample	# 3	0.182	21:59	
Breath Volume		1.152	21:59	
Air Blank		0.000	21:59	
Control Test		0.081	22:00	
Air Blank		0.000	22:00	
Diagnostics Check		OK	22:00	

Data Download Date/Time: 02/17/2016 14:53

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAIN,JEANETTE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: GAINES,JAMES
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 01/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1870563
DRY GAS EXPIRY DATE: 06/17/2018
VIOLATION CODE: DUI
ARREST TIME: 20:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:55

SUBJECT'S LAST NAME: STRAHL

DATE/TIME OF TEST: 01/23/2016 22:43:40

Diagnostics Check		OK	22:32	
Air Blank		0.000	22:32	
Control Test		0.081	22:33	
Air Blank		0.000	22:33	
Subject Sample	# 1	0.163	22:36	Volume Not Met
Breath Volume		0.019	22:36	
Air Blank		0.000	22:37	
Air Blank		0.000	22:39	
Subject Sample	# 2		22:42	No Sample Provided
Breath Volume		0.000	22:42	
Air Blank		0.000	22:42	
Control Test		0.080	22:43	
Air Blank		0.000	22:43	
Diagnostics Check		OK	22:43	

Data Download Date/Time: 02/17/2016 14:53

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAIN,JEANETTE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: GEE,JOHN
ARRESTING OFFICER AGENCY: Palm Springs PD
LAST AGENCY INSPECTION: 01/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1870563
DRY GAS EXPIRY DATE: 06/17/2018
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: ALTAMIRANO

DATE/TIME OF TEST: 01/24/2016 03:36:53

Diagnostics Check		OK	03:29
Air Blank		0.000	03:29
Control Test		0.082	03:30
Air Blank		0.000	03:30
Subject Sample	# 1	0.208	03:31
Breath Volume		1.242	03:31
Air Blank		0.000	03:32
Air Blank		0.000	03:34
Subject Sample	# 2	0.228	03:35
Breath Volume		2.398	03:35
Air Blank		0.000	03:35
Control Test		0.082	03:36
Air Blank		0.000	03:36
Diagnostics Check		OK	03:36

Data Download Date/Time: 02/17/2016 14:53

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CHIQUITO-RODRIGUEZ,LENNDY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 01/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1870563
DRY GAS EXPIRY DATE: 06/17/2018
VIOLATION CODE: DUI
ARREST TIME: 03:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:07

SUBJECT'S LAST NAME: CAMPOSECO-GUILLEN

DATE/TIME OF TEST: 01/24/2016 04:42:29

Diagnostics Check		OK	04:35
Air Blank		0.000	04:35
Control Test		0.082	04:36
Air Blank		0.000	04:36
Subject Sample	# 1	0.161	04:37
Breath Volume		2.785	04:37
Air Blank		0.000	04:38
Air Blank		0.000	04:39
Subject Sample	# 2	0.153	04:40
Breath Volume		2.628	04:40
Air Blank		0.000	04:41
Control Test		0.081	04:41
Air Blank		0.000	04:42
Diagnostics Check		OK	04:42

Data Download Date/Time: 02/17/2016 14:54

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARI,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: DAMERON,JUSTIN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 01/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1870563
DRY GAS EXPIRY DATE: 06/17/2018
VIOLATION CODE: DUI
ARREST TIME: 08:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:52

SUBJECT'S LAST NAME: RAMOS

DATE/TIME OF TEST: 01/24/2016 10:08:59

Diagnostics Check		OK	10:02
Air Blank		0.000	10:02
Control Test		0.083	10:03
Air Blank		0.000	10:03
Subject Sample	# 1	0.228	10:04
Breath Volume		1.761	10:04
Air Blank		0.000	10:04
Air Blank		0.000	10:06
Subject Sample	# 2	0.227	10:07
Breath Volume		1.441	10:07
Air Blank		0.000	10:08
Control Test		0.081	10:08
Air Blank		0.000	10:08
Diagnostics Check		OK	10:08

Data Download Date/Time: 02/17/2016 14:54

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARI,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: RIEGER,MITCHELL
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 01/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1870563
DRY GAS EXPIRY DATE: 06/17/2018
VIOLATION CODE: DUI
ARREST TIME: 12:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:41

SUBJECT'S LAST NAME: MATOS

DATE/TIME OF TEST: 01/24/2016 14:11:24

Diagnostics Check		OK	14:05
Air Blank		0.000	14:05
Control Test		0.083	14:06
Air Blank		0.000	14:06
Subject Sample	# 1	0.016	14:07
Breath Volume		1.308	14:07
Air Blank		0.000	14:07
Air Blank		0.000	14:09
Subject Sample	# 2	0.013	14:09
Breath Volume		1.132	14:09
Air Blank		0.000	14:10
Control Test		0.083	14:10
Air Blank		0.000	14:11
Diagnostics Check		OK	14:11

Data Download Date/Time: 02/17/2016 14:54

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FICARRA,CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 01/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1870563
DRY GAS EXPIRY DATE: 06/17/2018
VIOLATION CODE: DUI
ARREST TIME: 22:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: KINCAID

DATE/TIME OF TEST: 01/27/2016 00:34:02

Diagnostics Check		OK	00:27
Air Blank		0.000	00:28
Control Test		0.082	00:28
Air Blank		0.000	00:29
Subject Sample	# 1	0.000	00:29
Breath Volume		1.449	00:29
Air Blank		0.000	00:30
Air Blank		0.000	00:32
Subject Sample	# 2	0.000	00:32
Breath Volume		1.445	00:32
Air Blank		0.000	00:33
Control Test		0.082	00:33
Air Blank		0.000	00:33
Diagnostics Check		OK	00:34

Data Download Date/Time: 02/17/2016 14:54

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MARGOLIS,JOSEPH
ARRESTING OFFICER AGENCY: Lantana PD
LAST AGENCY INSPECTION: 01/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1870563
DRY GAS EXPIRY DATE: 06/17/2018
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:18

SUBJECT'S LAST NAME: LOPEZ-SANCHEZ

DATE/TIME OF TEST: 01/27/2016 23:54:52

Diagnostics Check		OK	23:48
Air Blank		0.000	23:49
Control Test		0.082	23:49
Air Blank		0.000	23:49
Subject Sample	# 1	0.158	23:50
Breath Volume		3.027	23:50
Air Blank		0.000	23:51
Air Blank		0.000	23:52
Subject Sample	# 2	0.152	23:53
Breath Volume		2.964	23:53
Air Blank		0.000	23:53
Control Test		0.081	23:54
Air Blank		0.000	23:54
Diagnostics Check		OK	23:54

Data Download Date/Time: 02/17/2016 14:54

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARI,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: HERNANDEZ,RAFAEL
ARRESTING OFFICER AGENCY: Palm Springs PD
LAST AGENCY INSPECTION: 01/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1870563
DRY GAS EXPIRY DATE: 06/17/2018
VIOLATION CODE: DUI
ARREST TIME: 13:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:29

SUBJECT'S LAST NAME: RUCCO

DATE/TIME OF TEST: 01/28/2016 14:20:30

Diagnostics Check		OK	14:12
Air Blank		0.000	14:13
Control Test		0.082	14:13
Air Blank		0.000	14:14
Subject Sample	# 1	0.126	14:15
Breath Volume		1.347	14:15
Air Blank		0.000	14:16
Air Blank		0.000	14:17
Subject Sample	# 2	0.113	14:18
Breath Volume		1.328	14:18
Air Blank		0.000	14:19
Control Test		0.081	14:19
Air Blank		0.000	14:20
Diagnostics Check		OK	14:20

Data Download Date/Time: 02/17/2016 14:54

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAIN,JEANETTE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MERCIER,DANIEL
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 01/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1870563
DRY GAS EXPIRY DATE: 06/17/2018
VIOLATION CODE: DUI
ARREST TIME: 23:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: BONILLA LOPEZ

DATE/TIME OF TEST: 01/29/2016 00:43:24

Diagnostics Check		OK	00:36
Air Blank		0.000	00:37
Control Test		0.081	00:37
Air Blank		0.000	00:38
Subject Sample	# 1	0.203	00:38
Breath Volume		2.343	00:38
Air Blank		0.000	00:39
Air Blank		0.000	00:41
Subject Sample	# 2	0.199	00:41
Breath Volume		2.531	00:41
Air Blank		0.000	00:42
Control Test		0.081	00:42
Air Blank		0.000	00:43
Diagnostics Check		OK	00:43

Data Download Date/Time: 02/17/2016 14:54

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FICARRA,CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 01/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1870563
DRY GAS EXPIRY DATE: 06/17/2018
VIOLATION CODE: DUI
ARREST TIME: 00:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: ATHIQUZZAMAN

DATE/TIME OF TEST: 01/29/2016 01:49:34

Diagnostics Check		OK	01:38	
Air Blank		0.000	01:38	
Control Test		0.082	01:38	
Air Blank		0.000	01:39	
Subject Sample	# 1		01:42	No Sample Provided
Breath Volume		0.000	01:42	
Air Blank		0.000	01:43	
Air Blank		0.000	01:45	
Subject Sample	# 2		01:48	No Sample Provided
Breath Volume		0.000	01:48	
Air Blank		0.000	01:48	
Control Test		0.081	01:48	
Air Blank		0.000	01:49	
Diagnostics Check		OK	01:49	

Data Download Date/Time: 02/17/2016 14:54

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FICARRA,CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 01/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1870563
DRY GAS EXPIRY DATE: 06/17/2018
VIOLATION CODE: DUI
ARREST TIME: 00:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:41

SUBJECT'S LAST NAME: ATHIQUZZAMAN

DATE/TIME OF TEST: 01/29/2016 02:11:36

Diagnostics Check		OK	02:05
Air Blank		0.000	02:05
Control Test		0.082	02:06
Air Blank		0.000	02:06
Subject Sample	# 1	0.184	02:07
Breath Volume		2.054	02:07
Air Blank		0.000	02:07
Air Blank		0.000	02:09
Subject Sample	# 2	0.190	02:10
Breath Volume		1.359	02:10
Air Blank		0.000	02:10
Control Test		0.082	02:11
Air Blank		0.000	02:11
Diagnostics Check		OK	02:11

Data Download Date/Time: 02/17/2016 14:54

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SCHNEIDER,JOSHUA
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 01/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1870563
DRY GAS EXPIRY DATE: 06/17/2018
VIOLATION CODE: DUI
ARREST TIME: 07:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:54

SUBJECT'S LAST NAME: PAPER

DATE/TIME OF TEST: 01/29/2016 08:26:51

Diagnostics Check		OK	08:20
Air Blank		0.000	08:20
Control Test		0.082	08:20
Air Blank		0.000	08:21
Subject Sample	# 1	0.000	08:22
Breath Volume		1.433	08:22
Air Blank		0.000	08:22
Air Blank		0.000	08:24
Subject Sample	# 2	0.000	08:25
Breath Volume		1.218	08:25
Air Blank		0.000	08:25
Control Test		0.082	08:26
Air Blank		0.000	08:26
Diagnostics Check		OK	08:26

Data Download Date/Time: 02/17/2016 14:54

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAIN,JEANETTE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FERNANDEZ,RYAN
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 01/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1870563
DRY GAS EXPIRY DATE: 06/17/2018
VIOLATION CODE: DUI
ARREST TIME: 17:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:40

SUBJECT'S LAST NAME: MCGUIRE

DATE/TIME OF TEST: 01/29/2016 18:15:41

Diagnostics Check		OK	18:06
Air Blank		0.000	18:06
Control Test		0.082	18:06
Air Blank		0.000	18:07
Subject Sample	# 1	0.212	18:09
Breath Volume		2.238	18:09
Air Blank		0.000	18:09
Air Blank		0.000	18:11
Subject Sample	# 2	0.208	18:13
Breath Volume		1.695	18:13
Air Blank		0.000	18:14
Control Test		0.081	18:15
Air Blank		0.000	18:15
Diagnostics Check		OK	18:15

Data Download Date/Time: 02/17/2016 14:54

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAIN,JEANETTE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: OWEN,MATTHEW
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 01/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1870563
DRY GAS EXPIRY DATE: 06/17/2018
VIOLATION CODE: DUI
ARREST TIME: 17:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:20

SUBJECT'S LAST NAME: CARISEO

DATE/TIME OF TEST: 01/29/2016 18:57:17

Diagnostics Check		OK	18:49
Air Blank		0.000	18:49
Control Test		0.081	18:49
Air Blank		0.000	18:50
Subject Sample	# 1	0.312	18:52
Breath Volume		2.015	18:52
Air Blank		0.000	18:53
Air Blank		0.000	18:54
Subject Sample	# 2	0.304	18:55
Breath Volume		1.406	18:55
Air Blank		0.000	18:56
Control Test		0.080	18:56
Air Blank		0.000	18:57
Diagnostics Check		OK	18:57

Data Download Date/Time: 02/17/2016 14:54

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARI,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: HANTON,MELINDA
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 01/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1870563
DRY GAS EXPIRY DATE: 06/17/2018
VIOLATION CODE: DUI
ARREST TIME: 02:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:32

SUBJECT'S LAST NAME: PADILLA

DATE/TIME OF TEST: 01/31/2016 03:12:06

Diagnostics Check		OK	03:05
Air Blank		0.000	03:06
Control Test		0.082	03:06
Air Blank		0.000	03:07
Subject Sample	# 1	0.185	03:07
Breath Volume		2.050	03:07
Air Blank		0.000	03:08
Air Blank		0.000	03:10
Subject Sample	# 2	0.186	03:10
Breath Volume		2.210	03:10
Air Blank		0.000	03:11
Control Test		0.082	03:11
Air Blank		0.000	03:12
Diagnostics Check		OK	03:12

Data Download Date/Time: 02/17/2016 14:54

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARI,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LYNCH,GREGORY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 01/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1870563
DRY GAS EXPIRY DATE: 06/17/2018
VIOLATION CODE: DUI
ARREST TIME: 04:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:55

SUBJECT'S LAST NAME: CASTILLO

DATE/TIME OF TEST: 01/31/2016 05:30:03

Diagnostics Check		OK	05:23
Air Blank		0.000	05:24
Control Test		0.082	05:24
Air Blank		0.000	05:24
Subject Sample	# 1	0.229	05:25
Breath Volume		2.703	05:25
Air Blank		0.000	05:26
Air Blank		0.000	05:28
Subject Sample	# 2	0.227	05:28
Breath Volume		2.781	05:28
Air Blank		0.000	05:29
Control Test		0.081	05:29
Air Blank		0.000	05:29
Diagnostics Check		OK	05:30

Data Download Date/Time: 02/17/2016 14:54

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARI,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FREY,JACOB
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 01/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1870563
DRY GAS EXPIRY DATE: 06/17/2018
VIOLATION CODE: DUI
ARREST TIME: 06:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:57

SUBJECT'S LAST NAME: CORREIA

DATE/TIME OF TEST: 01/31/2016 07:27:31

Diagnostics Check		OK	07:21
Air Blank		0.000	07:21
Control Test		0.081	07:22
Air Blank		0.000	07:22
Subject Sample	# 1	0.112	07:23
Breath Volume		2.140	07:23
Air Blank		0.000	07:23
Air Blank		0.000	07:25
Subject Sample	# 2	0.118	07:25
Breath Volume		2.800	07:25
Air Blank		0.000	07:26
Control Test		0.082	07:26
Air Blank		0.000	07:27
Diagnostics Check		OK	07:27

Data Download Date/Time: 02/17/2016 14:54

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARI,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SCHNEIDER,JOSHUA
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 01/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1870563
DRY GAS EXPIRY DATE: 06/17/2018
VIOLATION CODE: DUI
ARREST TIME: 14:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:28

SUBJECT'S LAST NAME: BRITO

DATE/TIME OF TEST: 01/31/2016 15:14:05

Diagnostics Check		OK	15:07
Air Blank		0.000	15:08
Control Test		0.082	15:08
Air Blank		0.000	15:08
Subject Sample	# 1	0.237	15:09
Breath Volume		2.425	15:09
Air Blank		0.000	15:10
Air Blank		0.000	15:12
Subject Sample	# 2	0.239	15:12
Breath Volume		2.335	15:12
Air Blank		0.000	15:13
Control Test		0.082	15:13
Air Blank		0.000	15:14
Diagnostics Check		OK	15:14

Data Download Date/Time: 02/17/2016 14:54

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006262

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESCOBAR,RUBEN,
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: IHNKEN,MATTHEW
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 12/01/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20314080AS
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 02:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: ZARRA

DATE/TIME OF TEST: 01/01/2016 03:15:25

Diagnostics Check		OK	03:08
Air Blank		0.000	03:09
Control Test		0.080	03:09
Air Blank		0.000	03:09
Subject Sample	# 1	0.223	03:10
Breath Volume		2.496	03:10
Air Blank		0.000	03:11
Air Blank		0.000	03:13
Subject Sample	# 2	0.220	03:13
Breath Volume		2.230	03:13
Air Blank		0.000	03:14
Control Test		0.079	03:14
Air Blank		0.000	03:15
Diagnostics Check		OK	03:15

Data Download Date/Time: 01/21/2016 15:09

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006262

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FOSTER,GRAEME,D
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: PEREZ,JOSH
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 12/01/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20314080AS
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 01:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: VILLEDA

DATE/TIME OF TEST: 01/12/2016 02:32:12

Diagnostics Check		OK	02:25
Air Blank		0.000	02:25
Control Test		0.081	02:25
Air Blank		0.000	02:26
Subject Sample	# 1	0.222	02:27
Breath Volume		2.207	02:27
Air Blank		0.000	02:28
Air Blank		0.000	02:30
Subject Sample	# 2	0.206	02:30
Breath Volume		3.710	02:30
Air Blank		0.000	02:31
Control Test		0.079	02:31
Air Blank		0.000	02:32
Diagnostics Check		OK	02:32

Data Download Date/Time: 01/21/2016 15:09

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006262

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FOSTER,GRAEME,D
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: RICHARDSON,JAYSON
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 12/01/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20314080AS
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: CASTRO

DATE/TIME OF TEST: 01/16/2016 02:38:32

Diagnostics Check		OK	02:32
Air Blank		0.000	02:32
Control Test		0.081	02:33
Air Blank		0.000	02:33
Subject Sample	# 1	0.164	02:33
Breath Volume		2.285	02:33
Air Blank		0.000	02:34
Air Blank		0.000	02:36
Subject Sample	# 2	0.165	02:36
Breath Volume		2.496	02:36
Air Blank		0.000	02:37
Control Test		0.080	02:37
Air Blank		0.000	02:38
Diagnostics Check		OK	02:38

Data Download Date/Time: 01/21/2016 15:09

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006262

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FROST,JOSEPH,W
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: WHITTET,JAMES
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 01/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20314080AS
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 23:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: JOHANSEN

DATE/TIME OF TEST: 01/30/2016 00:04:29

Diagnostics Check		OK	23:55	
Air Blank		0.000	23:56	
Control Test		0.080	23:56	
Air Blank		0.000	23:57	
Subject Sample	# 1		00:00	No Sample Provided
Breath Volume		0.000	00:00	
Air Blank		0.000	00:00	
Air Blank		0.000	00:02	
Subject Sample	# 2		00:03	Subject Test Refused
Breath Volume		0.000	00:03	
Air Blank		0.000	00:03	
Control Test		0.080	00:03	
Air Blank		0.000	00:04	
Diagnostics Check		OK	00:04	

Data Download Date/Time: 02/15/2016 12:03

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006262

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHEELER,TIMOTHY,R
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: WHEELER,TIMOTHY
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 01/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20314080AS
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: YAN

DATE/TIME OF TEST: 01/31/2016 03:40:21

Diagnostics Check		OK	03:32
Air Blank		0.000	03:32
Control Test		0.080	03:32
Air Blank		0.000	03:33
Subject Sample	# 1	0.065	03:35
Breath Volume		1.796	03:35
Air Blank		0.000	03:36
Air Blank		0.000	03:37
Subject Sample	# 2	0.062	03:38
Breath Volume		1.722	03:38
Air Blank		0.000	03:39
Control Test		0.078	03:39
Air Blank		0.000	03:40
Diagnostics Check		OK	03:40

Data Download Date/Time: 02/15/2016 12:03

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006262

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHEELER,TIMOTHY,R
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: SMITH,MICHELLE
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 01/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20314080AS
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: COPELAND

DATE/TIME OF TEST: 01/31/2016 04:18:48

Diagnostics Check		OK	04:15	
Air Blank		0.000	04:15	
Control Test		0.078	04:15	
Air Blank		0.000	04:16	
Subject Sample	# 1		04:17	Subject Test Refused
Breath Volume		0.000	04:17	
Air Blank		0.000	04:17	
Control Test		0.077	04:18	
Air Blank		0.000	04:18	
Diagnostics Check		OK	04:18	

Data Download Date/Time: 02/15/2016 12:03

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006384

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEMEN,AARON,M
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: STEMEN,AARON
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 12/16/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:26

SUBJECT'S LAST NAME: DOWLING

DATE/TIME OF TEST: 01/03/2016 02:53:13

Diagnostics Check		OK	02:46
Air Blank		0.000	02:47
Control Test		0.079	02:47
Air Blank		0.000	02:48
Subject Sample	# 1	0.102	02:48
Breath Volume		1.847	02:48
Air Blank		0.000	02:49
Air Blank		0.000	02:51
Subject Sample	# 2	0.111	02:51
Breath Volume		1.988	02:51
Air Blank		0.000	02:52
Control Test		0.079	02:52
Air Blank		0.000	02:53
Diagnostics Check		OK	02:53

Data Download Date/Time: 01/19/2016 21:24

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006384

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEMEN,AARON,M
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: STEMEN,AARON
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 12/16/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: FELIBERTI VAZQUEZ

DATE/TIME OF TEST: 01/08/2016 01:42:43

Diagnostics Check		OK	01:35
Air Blank		0.000	01:36
Control Test		0.080	01:36
Air Blank		0.000	01:36
Subject Sample	# 1	0.132	01:37
Breath Volume		2.144	01:37
Air Blank		0.000	01:38
Air Blank		0.000	01:40
Subject Sample	# 2	0.139	01:41
Breath Volume		2.000	01:41
Air Blank		0.000	01:41
Control Test		0.079	01:42
Air Blank		0.000	01:42
Diagnostics Check		OK	01:42

Data Download Date/Time: 01/19/2016 21:24

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006384

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEMEN,AARON,M
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: N/A,N/A
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 12/16/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: ZERO TOL
ARREST TIME: 00:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:32

SUBJECT'S LAST NAME: COMRIE

DATE/TIME OF TEST: 01/10/2016 01:04:36

Diagnostics Check		OK	00:52	
Air Blank		0.000	00:52	
Control Test		0.078	00:53	
Air Blank		0.000	00:53	
Subject Sample	# 1	0.023	00:56	Volume Not Met
Breath Volume		0.230	00:56	
Air Blank		0.000	00:57	
Air Blank		0.000	00:59	
Subject Sample	# 2	0.063	00:59	
Breath Volume		1.484	00:59	
Air Blank		0.000	01:00	
Air Blank		0.000	01:02	
Subject Sample	# 3	0.065	01:02	
Breath Volume		1.296	01:02	
Air Blank		0.000	01:03	
Control Test		0.079	01:03	
Air Blank		0.000	01:04	
Diagnostics Check		OK	01:04	

Data Download Date/Time: 01/19/2016 21:24

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006384

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEMEN,AARON,M
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: STEMEN,AARON
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 01/19/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:59

SUBJECT'S LAST NAME: BRAVO

DATE/TIME OF TEST: 01/22/2016 02:28:25

Diagnostics Check		OK	02:21
Air Blank		0.000	02:21
Control Test		0.079	02:22
Air Blank		0.000	02:22
Subject Sample	# 1	0.186	02:23
Breath Volume		2.328	02:23
Air Blank		0.000	02:24
Air Blank		0.000	02:26
Subject Sample	# 2	0.189	02:26
Breath Volume		2.781	02:26
Air Blank		0.000	02:27
Control Test		0.079	02:27
Air Blank		0.000	02:28
Diagnostics Check		OK	02:28

Data Download Date/Time: 02/13/2016 20:45

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006384

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITE,ROBERT,P
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: BRINSKO,THOMAS
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 01/19/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:18

SUBJECT'S LAST NAME: POLGAR

DATE/TIME OF TEST: 01/22/2016 04:45:18

Diagnostics Check		OK	04:39
Air Blank		0.000	04:39
Control Test		0.080	04:39
Air Blank		0.000	04:40
Subject Sample	# 1	0.000	04:41
Breath Volume		1.902	04:41
Air Blank		0.000	04:41
Air Blank		0.000	04:43
Subject Sample	# 2	0.000	04:43
Breath Volume		2.261	04:43
Air Blank		0.000	04:44
Control Test		0.077	04:44
Air Blank		0.000	04:45
Diagnostics Check		OK	04:45

Data Download Date/Time: 02/13/2016 20:45

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006384

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORENGO,LISA,
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: BALLENGER,CASANDRA
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 01/19/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 18:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:11

SUBJECT'S LAST NAME: EVANS

DATE/TIME OF TEST: 01/22/2016 19:43:18

Diagnostics Check		OK	19:31	
Air Blank		0.000	19:32	
Control Test		0.080	19:32	
Air Blank		0.000	19:33	
Subject Sample	# 1		19:36	No Sample Provided
Breath Volume		0.000	19:36	
Air Blank		0.000	19:36	
Air Blank		0.000	19:38	
Subject Sample	# 2		19:41	No Sample Provided
Breath Volume		0.000	19:41	
Air Blank		0.000	19:42	
Control Test		0.080	19:42	
Air Blank		0.000	19:43	
Diagnostics Check		OK	19:43	

Data Download Date/Time: 02/13/2016 20:45

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006384

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEMEN,AARON,M
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: STEMEN,AARON
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 01/19/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:03

SUBJECT'S LAST NAME: MARSH

DATE/TIME OF TEST: 01/28/2016 03:36:59

Diagnostics Check		OK	03:27	
Air Blank		0.000	03:27	
Control Test		0.079	03:28	
Air Blank		0.000	03:28	
Subject Sample	# 1		03:31	No Sample Provided
Breath Volume		0.000	03:31	
Air Blank		0.000	03:32	
Air Blank		0.000	03:34	
Subject Sample	# 2		03:35	Subject Test Refused
Breath Volume		0.000	03:35	
Air Blank		0.000	03:36	
Control Test		0.078	03:36	
Air Blank		0.000	03:36	
Diagnostics Check		OK	03:36	

Data Download Date/Time: 02/13/2016 20:45

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006384

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLLINS,MARVIN,D
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: COLLINS,MARVIN
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 01/19/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 14:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:29

SUBJECT'S LAST NAME: HARP

DATE/TIME OF TEST: 01/29/2016 15:02:13

Diagnostics Check		OK	14:50	
Air Blank		0.000	14:51	
Control Test		0.080	14:51	
Air Blank		0.000	14:52	
Subject Sample	# 1		14:55	No Sample Provided
Breath Volume		0.000	14:55	
Air Blank		0.000	14:55	
Air Blank		0.000	14:57	
Subject Sample	# 2		15:00	No Sample Provided
Breath Volume		0.000	15:00	
Air Blank		0.000	15:01	
Control Test		0.080	15:01	
Air Blank		0.000	15:02	
Diagnostics Check		OK	15:02	

Data Download Date/Time: 02/13/2016 20:45

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 12/10/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:03

SUBJECT'S LAST NAME: BRAGDON

DATE/TIME OF TEST: 01/01/2016 00:30:27

Diagnostics Check		OK	00:23
Air Blank		0.000	00:24
Control Test		0.081	00:24
Air Blank		0.000	00:25
Subject Sample	# 1	0.111	00:25
Breath Volume		2.605	00:25
Air Blank		0.000	00:26
Air Blank		0.000	00:28
Subject Sample	# 2	0.104	00:28
Breath Volume		2.324	00:28
Air Blank		0.000	00:29
Control Test		0.081	00:29
Air Blank		0.000	00:30
Diagnostics Check		OK	00:30

Data Download Date/Time: 01/04/2016 19:20

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY,PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ANGELICCHI,J
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/10/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:55

SUBJECT'S LAST NAME: MWERERU

DATE/TIME OF TEST: 01/01/2016 05:22:40

Diagnostics Check		OK	05:15
Air Blank		0.000	05:16
Control Test		0.081	05:16
Air Blank		0.000	05:17
Subject Sample	# 1	0.066	05:18
Breath Volume		1.691	05:18
Air Blank		0.000	05:18
Air Blank		0.000	05:20
Subject Sample	# 2	0.061	05:21
Breath Volume		2.035	05:21
Air Blank		0.000	05:21
Control Test		0.081	05:22
Air Blank		0.000	05:22
Diagnostics Check		OK	05:22

Data Download Date/Time: 01/04/2016 19:20

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: AGUAYO,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/10/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:26

SUBJECT'S LAST NAME: HOWARD

DATE/TIME OF TEST: 01/02/2016 05:53:34

Diagnostics Check		OK	05:46
Air Blank		0.000	05:47
Control Test		0.081	05:47
Air Blank		0.000	05:48
Subject Sample	# 1	0.138	05:48
Breath Volume		1.265	05:48
Air Blank		0.000	05:49
Air Blank		0.000	05:50
Subject Sample	# 2	0.125	05:51
Breath Volume		1.441	05:51
Air Blank		0.000	05:52
Control Test		0.082	05:52
Air Blank		0.000	05:53
Diagnostics Check		OK	05:53

Data Download Date/Time: 01/04/2016 19:20

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 12/10/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:26

SUBJECT'S LAST NAME: SUTTER JR

DATE/TIME OF TEST: 01/03/2016 02:53:01

Diagnostics Check		OK	02:46
Air Blank		0.000	02:46
Control Test		0.080	02:46
Air Blank		0.000	02:47
Subject Sample	# 1	0.179	02:48
Breath Volume		4.035	02:48
Air Blank		0.000	02:48
Air Blank		0.000	02:50
Subject Sample	# 2	0.182	02:51
Breath Volume		3.910	02:51
Air Blank		0.000	02:52
Control Test		0.080	02:52
Air Blank		0.000	02:52
Diagnostics Check		OK	02:53

Data Download Date/Time: 01/04/2016 19:20

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK,CASEY,S
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WOODEN,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/10/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:27

SUBJECT'S LAST NAME: KNAUER

DATE/TIME OF TEST: 01/04/2016 00:55:04

Diagnostics Check		OK	00:48
Air Blank		0.000	00:49
Control Test		0.081	00:49
Air Blank		0.000	00:49
Subject Sample	# 1	0.133	00:50
Breath Volume		3.175	00:50
Air Blank		0.000	00:51
Air Blank		0.000	00:52
Subject Sample	# 2	0.131	00:53
Breath Volume		3.648	00:53
Air Blank		0.000	00:54
Control Test		0.081	00:54
Air Blank		0.000	00:54
Diagnostics Check		OK	00:55

Data Download Date/Time: 01/04/2016 19:20

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SIMPKINS,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/10/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:17

SUBJECT'S LAST NAME: DOTSEY

DATE/TIME OF TEST: 01/04/2016 01:48:57

Diagnostics Check		OK	01:37	
Air Blank		0.000	01:37	
Control Test		0.081	01:38	
Air Blank		0.000	01:38	
Subject Sample	# 1		01:41	No Sample Provided
Breath Volume		0.000	01:41	
Air Blank		0.000	01:42	
Air Blank		0.000	01:44	
Subject Sample	# 2		01:47	No Sample Provided
Breath Volume		0.000	01:47	
Air Blank		0.000	01:48	
Control Test		0.082	01:48	
Air Blank		0.000	01:48	
Diagnostics Check		OK	01:48	

Data Download Date/Time: 01/04/2016 19:20

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006386

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPBELL,ADAM,M
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: CAMPBELL,ADAM
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: FUENTES

DATE/TIME OF TEST: 01/09/2016 01:04:31

Diagnostics Check		OK	01:01	
Air Blank		0.000	01:01	
Control Test		0.076	01:01	
Air Blank		0.000	01:02	
Subject Sample	# 1		01:02	Subject Test Refused
Breath Volume		0.000	01:02	
Air Blank		0.000	01:03	
Control Test		0.078	01:03	
Air Blank		0.000	01:04	
Diagnostics Check		OK	01:04	

Data Download Date/Time: 01/29/2016 13:34

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006386

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPBELL,ADAM,M
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: TYLER,CHRISTOPHER
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: ROMAN MARRERO

DATE/TIME OF TEST: 01/29/2016 00:17:31

Diagnostics Check	OK	00:15	
Air Blank	0.000	00:16	
Control Test	0.074	00:16	Control Outside Tolerance
Air Blank	0.000	00:17	

Data Download Date/Time: 01/29/2016 13:34

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006386

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPBELL,ADAM,M
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: TYLER,CHRISTOPHER
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: ROMAN MARRERO

DATE/TIME OF TEST: 01/29/2016 00:25:39

Diagnostics Check		OK	00:18
Air Blank		0.000	00:19
Control Test		0.077	00:19
Air Blank		0.000	00:20
Subject Sample	# 1	0.144	00:20
Breath Volume		3.171	00:20
Air Blank		0.000	00:21
Air Blank		0.000	00:23
Subject Sample	# 2	0.135	00:23
Breath Volume		1.484	00:23
Air Blank		0.000	00:24
Control Test		0.076	00:24
Air Blank		0.000	00:25
Diagnostics Check		OK	00:25

Data Download Date/Time: 01/29/2016 13:34

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STCLAIR,TODD,C
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: STCLAIR,TODD
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 12/29/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 18:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:39

SUBJECT'S LAST NAME: BLACKBURN

DATE/TIME OF TEST: 01/05/2016 19:14:22

Diagnostics Check		OK	19:07
Air Blank		0.000	19:08
Control Test		0.081	19:08
Air Blank		0.000	19:08
Subject Sample	# 1	0.244	19:09
Breath Volume		2.800	19:09
Air Blank		0.000	19:10
Air Blank		0.000	19:12
Subject Sample	# 2	0.234	19:12
Breath Volume		2.207	19:12
Air Blank		0.000	19:13
Control Test		0.077	19:13
Air Blank		0.000	19:14
Diagnostics Check		OK	19:14

Data Download Date/Time: 01/26/2016 09:33

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TURNER,MARK,A
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: TURNER,MARK
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 12/29/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: EVERETTE

DATE/TIME OF TEST: 01/08/2016 03:36:53

Diagnostics Check		OK	03:28	
Air Blank		0.000	03:29	
Control Test		0.081	03:29	
Air Blank		0.000	03:29	
Subject Sample	# 1		03:33	No Sample Provided
Breath Volume		0.000	03:33	
Air Blank		0.000	03:33	
Air Blank		0.000	03:35	
Subject Sample	# 2		03:36	Sequence Aborted
Breath Volume		0.000	03:36	
Air Blank		0.000	03:36	

Data Download Date/Time: 01/26/2016 09:33

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TUCKER,JOSHUA,M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: TUCKER,JOSHUA
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 12/29/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:05

SUBJECT'S LAST NAME: SASSE

DATE/TIME OF TEST: 01/10/2016 21:36:14

Diagnostics Check		OK	21:32	
Air Blank		0.000	21:32	
Control Test		0.081	21:32	
Air Blank		0.000	21:33	
Subject Sample	# 1		21:34	Subject Test Refused
Breath Volume		0.000	21:34	
Air Blank		0.000	21:35	
Control Test		0.082	21:35	
Air Blank		0.000	21:36	
Diagnostics Check		OK	21:36	

Data Download Date/Time: 01/26/2016 09:33

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BEVIS,T,D
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BEVIS,T
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 12/29/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 12:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 01/14/2016 01:23:08

Diagnostics Check		OK	01:16	
Air Blank		0.000	01:17	
Control Test		0.080	01:17	
Air Blank		0.000	01:18	
Subject Sample	# 1	0.182	01:18	
Breath Volume		1.761	01:18	
Air Blank		0.000	01:19	
Air Blank		0.000	01:21	
Subject Sample	# 2	0.168	01:21	
Breath Volume		1.191	01:21	
Air Blank		0.000	01:22	
Control Test		0.060	01:22	Control Outside Tolerance
Air Blank		0.000	01:23	

Data Download Date/Time: 01/26/2016 09:33

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROOP,CHAD,E
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: SHORETTE,BRIAN
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 12/29/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:06

SUBJECT'S LAST NAME: NEVISON

DATE/TIME OF TEST: 01/14/2016 02:33:14

Diagnostics Check		OK	02:26
Air Blank		0.000	02:27
Control Test		0.079	02:27
Air Blank		0.000	02:27
Subject Sample	# 1	0.181	02:28
Breath Volume		2.140	02:28
Air Blank		0.000	02:29
Air Blank		0.000	02:31
Subject Sample	# 2	0.176	02:31
Breath Volume		2.425	02:31
Air Blank		0.000	02:32
Control Test		0.080	02:32
Air Blank		0.000	02:33
Diagnostics Check		OK	02:33

Data Download Date/Time: 01/26/2016 09:33

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BEVIS,T,D
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BEVIS,T
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 12/29/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 10:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: BLAKE

DATE/TIME OF TEST: 01/14/2016 23:08:05

Diagnostics Check		OK	23:02
Air Blank		0.000	23:02
Control Test		0.080	23:02
Air Blank		0.000	23:03
Subject Sample	# 1	0.242	23:03
Breath Volume		1.523	23:03
Air Blank		0.000	23:04
Air Blank		0.000	23:06
Subject Sample	# 2	0.242	23:06
Breath Volume		2.324	23:06
Air Blank		0.000	23:07
Control Test		0.079	23:07
Air Blank		0.000	23:07
Diagnostics Check		OK	23:08

Data Download Date/Time: 01/26/2016 09:33

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BEVIS,T,D
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BEVIS,T
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 12/29/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:43

SUBJECT'S LAST NAME: REEVES

DATE/TIME OF TEST: 01/16/2016 05:21:56

Diagnostics Check		OK	05:16	
Air Blank		0.000	05:17	
Control Test		0.081	05:17	
Air Blank		0.000	05:17	
Subject Sample	# 1	0.159	05:18	
Breath Volume		1.742	05:18	
Air Blank		0.000	05:19	
Air Blank			05:21	Ambient Fail
Air Blank		0.000	05:21	

Data Download Date/Time: 01/26/2016 09:33

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BEVIS,T,D
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BEVIS,T
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 12/29/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:43

SUBJECT'S LAST NAME: REEVES

DATE/TIME OF TEST: 01/16/2016 05:30:21

Diagnostics Check		OK	05:23
Air Blank		0.000	05:23
Control Test		0.080	05:24
Air Blank		0.000	05:24
Subject Sample	# 1	0.165	05:25
Breath Volume		1.617	05:25
Air Blank		0.000	05:26
Air Blank		0.000	05:27
Subject Sample	# 2	0.157	05:28
Breath Volume		2.054	05:28
Air Blank		0.000	05:29
Control Test		0.080	05:29
Air Blank		0.000	05:30
Diagnostics Check		OK	05:30

Data Download Date/Time: 01/26/2016 09:33

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROOP,CHAD,E
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: ROOP,CHAD
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 12/29/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:51

SUBJECT'S LAST NAME: HALL

DATE/TIME OF TEST: 01/17/2016 00:24:24

Diagnostics Check		OK	00:19	
Air Blank		0.000	00:20	
Control Test		0.081	00:20	
Air Blank		0.000	00:21	
Subject Sample	# 1		00:23	Subject Test Refused
Breath Volume		0.000	00:23	
Air Blank		0.000	00:23	
Control Test		0.082	00:23	
Air Blank		0.000	00:24	
Diagnostics Check		OK	00:24	

Data Download Date/Time: 01/26/2016 09:33

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BEVIS,T,D
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BEVIS,T
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 12/29/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:11

SUBJECT'S LAST NAME: POPPE

DATE/TIME OF TEST: 01/17/2016 03:48:38

Diagnostics Check		OK	03:41
Air Blank		0.000	03:42
Control Test		0.081	03:42
Air Blank		0.000	03:43
Subject Sample	# 1	0.068	03:44
Breath Volume		3.527	03:44
Air Blank		0.000	03:44
Air Blank		0.000	03:46
Subject Sample	# 2	0.062	03:47
Breath Volume		2.285	03:47
Air Blank		0.000	03:47
Control Test		0.081	03:48
Air Blank		0.000	03:48
Diagnostics Check		OK	03:48

Data Download Date/Time: 01/26/2016 09:33

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BEVIS,T,D
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BEVIS,T
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 12/29/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:47

SUBJECT'S LAST NAME: PENTON

DATE/TIME OF TEST: 01/22/2016 23:32:59

Diagnostics Check	OK	23:31	
Air Blank	0.000	23:32	
Control Test	0.056	23:32	Control Outside Tolerance
Air Blank	0.000	23:32	

Data Download Date/Time: 01/26/2016 09:33

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BEVIS,T,D
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BEVIS,T
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 12/29/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:47

SUBJECT'S LAST NAME: PENTON

DATE/TIME OF TEST: 01/22/2016 23:35:52

Diagnostics Check	OK	23:34	
Air Blank		23:35	Ambient Fail
Air Blank	0.000	23:35	

Data Download Date/Time: 01/26/2016 09:33

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BEVIS,T,D
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BEVIS,T
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 12/29/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:47

SUBJECT'S LAST NAME: PENTON

DATE/TIME OF TEST: 01/22/2016 23:40:04

Diagnostics Check	OK	23:38	
Air Blank	0.000	23:39	
Control Test	0.073	23:39	Control Outside Tolerance
Air Blank	0.000	23:40	

Data Download Date/Time: 01/26/2016 09:33

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHORETTE,BRIAN,E
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: SHORETTE,BRIAN
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 12/29/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:54

SUBJECT'S LAST NAME: HENDERSON

DATE/TIME OF TEST: 01/23/2016 22:21:30

Diagnostics Check		OK	22:14
Air Blank		0.000	22:15
Control Test		0.081	22:15
Air Blank		0.000	22:16
Subject Sample	# 1	0.073	22:16
Breath Volume		1.417	22:16
Air Blank		0.000	22:17
Air Blank		0.000	22:19
Subject Sample	# 2	0.082	22:19
Breath Volume		1.484	22:19
Air Blank		0.000	22:20
Control Test		0.081	22:20
Air Blank		0.000	22:21
Diagnostics Check		OK	22:21

Data Download Date/Time: 01/26/2016 09:33

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TODD,DAVID,P
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: TODD,DAVID
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 12/29/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 16:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:50

SUBJECT'S LAST NAME: DUNLAVY

DATE/TIME OF TEST: 01/25/2016 17:16:51

Diagnostics Check		OK	17:10
Air Blank		0.000	17:11
Control Test		0.080	17:11
Air Blank		0.000	17:12
Subject Sample	# 1	0.142	17:12
Breath Volume		3.886	17:12
Air Blank		0.000	17:13
Air Blank		0.000	17:14
Subject Sample	# 2	0.140	17:15
Breath Volume		4.304	17:15
Air Blank		0.000	17:15
Control Test		0.080	17:16
Air Blank		0.000	17:16
Diagnostics Check		OK	17:16

Data Download Date/Time: 01/26/2016 09:33

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEVENS,MITCHELL,M
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: STEVENS,MITCHELL
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 12/29/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:54

SUBJECT'S LAST NAME: HOFER

DATE/TIME OF TEST: 01/01/2016 00:21:36

Diagnostics Check		OK	00:14
Air Blank		0.000	00:15
Control Test		0.081	00:15
Air Blank		0.000	00:16
Subject Sample	# 1	0.144	00:16
Breath Volume		2.585	00:16
Air Blank		0.000	00:17
Air Blank		0.000	00:19
Subject Sample	# 2	0.138	00:19
Breath Volume		2.031	00:19
Air Blank		0.000	00:20
Control Test		0.079	00:20
Air Blank		0.000	00:21
Diagnostics Check		OK	00:21

Data Download Date/Time: 01/26/2016 09:30

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUDSON,JOSHUA,K
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: HUDSON,JOSHUA
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 12/29/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:38

SUBJECT'S LAST NAME: BAILEY

DATE/TIME OF TEST: 01/01/2016 01:16:44

Diagnostics Check		OK	01:10
Air Blank		0.000	01:10
Control Test		0.081	01:10
Air Blank		0.000	01:11
Subject Sample	# 1	0.000	01:12
Breath Volume		1.914	01:12
Air Blank		0.000	01:12
Air Blank		0.000	01:14
Subject Sample	# 2	0.000	01:15
Breath Volume		1.640	01:15
Air Blank		0.000	01:15
Control Test		0.081	01:16
Air Blank		0.000	01:16
Diagnostics Check		OK	01:16

Data Download Date/Time: 01/26/2016 09:30

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROOP,CHAD,E
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: ROOP,CHAD
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 12/29/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 06:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:10

SUBJECT'S LAST NAME: PHILLIPS

DATE/TIME OF TEST: 01/02/2016 07:42:28

Diagnostics Check		OK	07:31	
Air Blank		0.000	07:31	
Control Test		0.082	07:31	
Air Blank		0.000	07:32	
Subject Sample	# 1		07:35	No Sample Provided
Breath Volume		0.000	07:35	
Air Blank		0.000	07:35	
Air Blank		0.000	07:37	
Subject Sample	# 2		07:41	No Sample Provided
Breath Volume		0.000	07:41	
Air Blank		0.000	07:41	
Control Test		0.082	07:41	
Air Blank		0.000	07:42	
Diagnostics Check		OK	07:42	

Data Download Date/Time: 01/26/2016 09:30

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROOP,CHAD,E
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: ROOP,CHAD
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 12/29/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 06:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:40

SUBJECT'S LAST NAME: PHILLIPS

DATE/TIME OF TEST: 01/02/2016 08:06:42

Diagnostics Check		OK	08:00	
Air Blank		0.000	08:01	
Control Test		0.083	08:01	
Air Blank		0.000	08:01	
Subject Sample	# 1	0.153	08:02	
Breath Volume		1.894	08:02	
Air Blank		0.000	08:03	
Air Blank		0.000	08:04	
Subject Sample	# 2		08:05	Subject Test Refused
Breath Volume		0.000	08:05	
Air Blank		0.000	08:05	
Control Test		0.081	08:06	
Air Blank		0.000	08:06	
Diagnostics Check		OK	08:06	

Data Download Date/Time: 01/26/2016 09:30

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WRIGHT,L,
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: WRIGHT,L
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 12/29/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:33

SUBJECT'S LAST NAME: PHAN

DATE/TIME OF TEST: 01/04/2016 01:05:43

Diagnostics Check		OK	01:01	
Air Blank		0.000	01:01	
Control Test		0.081	01:01	
Air Blank		0.000	01:02	
Subject Sample	# 1		01:04	Subject Test Refused
Breath Volume		0.175	01:04	
Air Blank		0.000	01:04	
Control Test		0.082	01:05	
Air Blank		0.000	01:05	
Diagnostics Check		OK	01:05	

Data Download Date/Time: 01/26/2016 09:30

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHELLEY,SAMUEL,L
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: SHELLEY,SAMUEL
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 12/29/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 17:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:15

SUBJECT'S LAST NAME: MITCHEM

DATE/TIME OF TEST: 01/07/2016 18:42:11

Diagnostics Check		OK	18:35
Air Blank		0.000	18:35
Control Test		0.080	18:36
Air Blank		0.000	18:36
Subject Sample	# 1	0.178	18:37
Breath Volume		2.566	18:37
Air Blank		0.000	18:38
Air Blank		0.000	18:39
Subject Sample	# 2	0.173	18:40
Breath Volume		2.511	18:40
Air Blank		0.000	18:41
Control Test		0.081	18:41
Air Blank		0.000	18:42
Diagnostics Check		OK	18:42

Data Download Date/Time: 01/26/2016 09:36

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROOP,CHAD,E
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: ROOP,CHAD
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 12/29/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: CURRAN

DATE/TIME OF TEST: 01/07/2016 23:11:01

Diagnostics Check		OK	23:08	
Air Blank		0.000	23:08	
Control Test		0.082	23:08	
Air Blank		0.000	23:09	
Subject Sample	# 1		23:09	Subject Test Refused
Breath Volume		0.000	23:09	
Air Blank		0.000	23:10	
Control Test		0.082	23:10	
Air Blank		0.000	23:10	
Diagnostics Check		OK	23:11	

Data Download Date/Time: 01/26/2016 09:36

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POOL,HAROLD,M
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: POOL,HAROLD
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 12/29/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: WOLTZ

DATE/TIME OF TEST: 01/08/2016 03:06:53

Diagnostics Check		OK	03:00
Air Blank		0.000	03:00
Control Test		0.081	03:01
Air Blank		0.000	03:01
Subject Sample	# 1	0.000	03:02
Breath Volume		1.445	03:02
Air Blank		0.000	03:02
Air Blank		0.000	03:04
Subject Sample	# 2	0.000	03:05
Breath Volume		1.203	03:05
Air Blank		0.000	03:05
Control Test		0.081	03:06
Air Blank		0.000	03:06
Diagnostics Check		OK	03:06

Data Download Date/Time: 01/26/2016 09:36

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROOP,CHAD,E
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: ROOP,CHAD
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 12/29/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:02

SUBJECT'S LAST NAME: TRAN

DATE/TIME OF TEST: 01/09/2016 02:25:18

Diagnostics Check		OK	02:22	
Air Blank		0.000	02:22	
Control Test		0.081	02:23	
Air Blank		0.000	02:23	
Subject Sample	# 1		02:23	Subject Test Refused
Breath Volume		0.000	02:23	
Air Blank		0.000	02:24	
Control Test		0.081	02:24	
Air Blank		0.000	02:25	
Diagnostics Check		OK	02:25	

Data Download Date/Time: 01/26/2016 09:36

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUDSON,JOSHUA,K
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: HUDSON,JOSHUA
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 12/29/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:24

SUBJECT'S LAST NAME: COURSEY

DATE/TIME OF TEST: 01/09/2016 23:12:01

Diagnostics Check		OK	23:00	
Air Blank		0.000	23:01	
Control Test		0.080	23:01	
Air Blank		0.000	23:01	
Subject Sample	# 1		23:05	No Sample Provided
Breath Volume		0.000	23:05	
Air Blank		0.000	23:05	
Air Blank		0.000	23:07	
Subject Sample	# 2		23:10	No Sample Provided
Breath Volume		0.000	23:10	
Air Blank		0.000	23:11	
Control Test		0.081	23:11	
Air Blank		0.000	23:11	
Diagnostics Check		OK	23:12	

Data Download Date/Time: 01/26/2016 09:36

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUDSON,JOSHUA,K
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: HUDSON,JOSHUA
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 12/29/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:48

SUBJECT'S LAST NAME: GIBBS

DATE/TIME OF TEST: 01/10/2016 02:15:31

Diagnostics Check		OK	02:12	
Air Blank		0.000	02:12	
Control Test		0.080	02:13	
Air Blank		0.000	02:13	
Subject Sample	# 1		02:14	RFI Detect
Breath Volume		0.000	02:14	
Air Blank		0.000	02:14	
Control Test		0.081	02:14	
Air Blank		0.000	02:15	
Diagnostics Check		OK	02:15	

Data Download Date/Time: 01/26/2016 09:36

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUDSON,JOSHUA,K
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: HUDSON,JOSHUA
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 12/29/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:48

SUBJECT'S LAST NAME: GIBBS

DATE/TIME OF TEST: 01/10/2016 02:24:37

Diagnostics Check		OK	02:17
Air Blank		0.000	02:17
Control Test		0.080	02:17
Air Blank		0.000	02:18
Subject Sample	# 1	0.202	02:19
Breath Volume		1.148	02:19
Air Blank		0.000	02:20
Air Blank		0.000	02:22
Subject Sample	# 2	0.199	02:22
Breath Volume		2.746	02:22
Air Blank		0.000	02:23
Control Test		0.079	02:23
Air Blank		0.000	02:24
Diagnostics Check		OK	02:24

Data Download Date/Time: 01/26/2016 09:36

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REGISTER,MICHAEL,A
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: REGISTER,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 12/29/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: STANLEY

DATE/TIME OF TEST: 01/10/2016 03:23:46

Diagnostics Check		OK	03:20	
Air Blank		0.000	03:21	
Control Test		0.080	03:21	
Air Blank		0.000	03:22	
Subject Sample	# 1		03:22	Subject Test Refused
Breath Volume		0.000	03:22	
Air Blank		0.000	03:22	
Control Test		0.080	03:23	
Air Blank		0.000	03:23	
Diagnostics Check		OK	03:23	

Data Download Date/Time: 01/26/2016 09:36

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUDSON,JOSHUA,K
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: HUDSON,JOSHUA
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 12/29/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:54

SUBJECT'S LAST NAME: CURTIS

DATE/TIME OF TEST: 01/10/2016 21:31:09

Diagnostics Check		OK	21:19	
Air Blank		0.000	21:20	
Control Test		0.081	21:20	
Air Blank		0.000	21:21	
Subject Sample	# 1		21:24	No Sample Provided
Breath Volume		0.000	21:24	
Air Blank		0.000	21:24	
Air Blank		0.000	21:26	
Subject Sample	# 2		21:29	No Sample Provided
Breath Volume		0.000	21:29	
Air Blank		0.000	21:30	
Control Test		0.081	21:30	
Air Blank		0.000	21:31	
Diagnostics Check		OK	21:31	

Data Download Date/Time: 01/26/2016 09:36

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BEARD,JOHN,K
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: BEARD,JOHN
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 12/29/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:56

SUBJECT'S LAST NAME: CADEM

DATE/TIME OF TEST: 01/12/2016 05:25:08

Diagnostics Check		OK	05:17
Air Blank		0.000	05:17
Control Test		0.082	05:18
Air Blank		0.000	05:18
Subject Sample	# 1	0.011	05:19
Breath Volume		2.046	05:19
Air Blank		0.000	05:20
Air Blank		0.000	05:22
Subject Sample	# 2	0.012	05:23
Breath Volume		3.097	05:23
Air Blank		0.000	05:24
Control Test		0.082	05:24
Air Blank		0.000	05:25
Diagnostics Check		OK	05:25

Data Download Date/Time: 01/26/2016 09:36

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEE,MIKEL,A
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: LEE,MIKEL
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 12/29/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:54

SUBJECT'S LAST NAME: PENNINGTON

DATE/TIME OF TEST: 01/14/2016 01:26:45

Diagnostics Check		OK	01:15	
Air Blank		0.000	01:15	
Control Test		0.082	01:15	
Air Blank		0.000	01:16	
Subject Sample	# 1		01:19	No Sample Provided
Breath Volume		0.000	01:19	
Air Blank		0.000	01:19	
Air Blank		0.000	01:21	
Subject Sample	# 2	0.084	01:25	Volume Not Met
Breath Volume		0.820	01:25	
Air Blank		0.000	01:25	
Control Test		0.081	01:26	
Air Blank		0.000	01:26	
Diagnostics Check		OK	01:26	

Data Download Date/Time: 01/26/2016 09:36

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEE,MIKEL,A
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: LEE,MIKEL
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 12/29/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: PENNINGTON

DATE/TIME OF TEST: 01/14/2016 01:52:45

Diagnostics Check		OK	01:45
Air Blank		0.000	01:46
Control Test		0.080	01:46
Air Blank		0.000	01:46
Subject Sample	# 1	0.089	01:47
Breath Volume		1.117	01:47
Air Blank		0.000	01:48
Air Blank		0.000	01:50
Subject Sample	# 2	0.086	01:51
Breath Volume		1.230	01:51
Air Blank		0.000	01:51
Control Test		0.080	01:52
Air Blank		0.000	01:52
Diagnostics Check		OK	01:52

Data Download Date/Time: 01/26/2016 09:36

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROOP,CHAD,E
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: ROOP,CHAD
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 12/29/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: LOCKHART

DATE/TIME OF TEST: 01/15/2016 04:26:47

Diagnostics Check		OK	04:20
Air Blank		0.000	04:20
Control Test		0.081	04:20
Air Blank		0.000	04:21
Subject Sample	# 1	0.158	04:22
Breath Volume		3.203	04:22
Air Blank		0.000	04:22
Air Blank		0.000	04:24
Subject Sample	# 2	0.149	04:25
Breath Volume		2.554	04:25
Air Blank		0.000	04:25
Control Test		0.080	04:26
Air Blank		0.000	04:26
Diagnostics Check		OK	04:26

Data Download Date/Time: 01/26/2016 09:36

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROOP,CHAD,E
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: ROOP,CHAD
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 12/29/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:51

SUBJECT'S LAST NAME: HALL

DATE/TIME OF TEST: 01/17/2016 00:13:35

Diagnostics Check	OK	00:12	
Air Blank	0.000	00:12	
Control Test	0.095	00:12	Control Outside Tolerance
Air Blank	0.000	00:13	

Data Download Date/Time: 01/26/2016 09:36

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POOL,HAROLD,N
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: TEST,TEST
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 12/29/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: OTHER
ARREST TIME: 23:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: TEST

DATE/TIME OF TEST: 01/17/2016 00:29:59

Diagnostics Check		OK	00:24
Air Blank		0.000	00:24
Control Test		0.080	00:25
Air Blank		0.000	00:25
Subject Sample	# 1	0.000	00:25
Breath Volume		2.187	00:25
Air Blank		0.000	00:26
Air Blank		0.000	00:28
Subject Sample	# 2	0.000	00:28
Breath Volume		2.253	00:28
Air Blank		0.000	00:29
Control Test		0.080	00:29
Air Blank		0.000	00:29
Diagnostics Check		OK	00:29

Data Download Date/Time: 01/26/2016 09:36

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEEKS,JEREMIAH,S
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: MEEKS,JEREMIAH
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 12/29/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: WRISLEY

DATE/TIME OF TEST: 01/17/2016 04:02:46

Diagnostics Check		OK	03:55
Air Blank		0.000	03:55
Control Test		0.081	03:56
Air Blank		0.000	03:56
Subject Sample	# 1	0.219	03:57
Breath Volume		1.835	03:57
Air Blank		0.000	03:58
Air Blank		0.000	03:59
Subject Sample	# 2	0.220	04:01
Breath Volume		1.808	04:01
Air Blank		0.000	04:01
Control Test		0.077	04:02
Air Blank		0.000	04:02
Diagnostics Check		OK	04:02

Data Download Date/Time: 01/26/2016 09:36

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROOP,CHAD,E
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: ROOP,CHAD
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 12/29/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:27

SUBJECT'S LAST NAME: NISBETT

DATE/TIME OF TEST: 01/22/2016 02:54:42

Diagnostics Check		OK	02:48
Air Blank		0.000	02:48
Control Test		0.080	02:48
Air Blank		0.000	02:49
Subject Sample	# 1	0.099	02:50
Breath Volume		2.031	02:50
Air Blank		0.000	02:50
Air Blank		0.000	02:52
Subject Sample	# 2	0.098	02:53
Breath Volume		1.300	02:53
Air Blank		0.000	02:53
Control Test		0.080	02:54
Air Blank		0.000	02:54
Diagnostics Check		OK	02:54

Data Download Date/Time: 01/26/2016 09:36

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUDSON,JOSHUA,K
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: HUDSON,JOSHUA
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 12/29/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 17:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:30

SUBJECT'S LAST NAME: BOLING

DATE/TIME OF TEST: 01/22/2016 18:07:24

Diagnostics Check		OK	17:55	
Air Blank		0.000	17:56	
Control Test		0.081	17:56	
Air Blank		0.000	17:57	
Subject Sample	# 1		18:00	No Sample Provided
Breath Volume		0.000	18:00	
Air Blank		0.000	18:00	
Air Blank		0.000	18:02	
Subject Sample	# 2		18:06	No Sample Provided
Breath Volume		0.000	18:06	
Air Blank		0.000	18:06	
Control Test		0.081	18:06	
Air Blank		0.000	18:07	
Diagnostics Check		OK	18:07	

Data Download Date/Time: 01/26/2016 09:36

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BEVIS,T,D
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BEVIS,T
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 12/29/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:47

SUBJECT'S LAST NAME: PENTON

DATE/TIME OF TEST: 01/22/2016 23:50:55

Diagnostics Check		OK	23:43
Air Blank		0.000	23:43
Control Test		0.081	23:44
Air Blank		0.000	23:44
Subject Sample	# 1	0.159	23:45
Breath Volume		1.585	23:45
Air Blank		0.000	23:46
Air Blank		0.000	23:47
Subject Sample	# 2	0.159	23:49
Breath Volume		1.605	23:49
Air Blank		0.000	23:49
Control Test		0.081	23:50
Air Blank		0.000	23:50
Diagnostics Check		OK	23:50

Data Download Date/Time: 01/26/2016 09:36

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAYO,TREVOR,W
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: MAYO,TREVOR
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 12/29/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: RAJAH

DATE/TIME OF TEST: 01/23/2016 01:56:46

Diagnostics Check		OK	01:51	
Air Blank		0.000	01:51	
Control Test		0.081	01:52	
Air Blank		0.000	01:52	
Subject Sample	# 1		01:55	RFI Detect
Breath Volume		0.000	01:55	
Air Blank		0.000	01:55	
Control Test		0.081	01:56	
Air Blank		0.000	01:56	
Diagnostics Check		OK	01:56	

Data Download Date/Time: 01/26/2016 09:36

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAYO,TREVOR,W
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: MAYO,TREVOR
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 12/29/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: RAJAH

DATE/TIME OF TEST: 01/23/2016 02:09:44

Diagnostics Check		OK	01:58	
Air Blank		0.000	01:58	
Control Test		0.080	01:58	
Air Blank		0.000	01:59	
Subject Sample	# 1	0.131	02:02	Volume Not Met
Breath Volume		0.167	02:02	
Air Blank		0.000	02:03	
Air Blank		0.000	02:04	
Subject Sample	# 2	0.144	02:08	Volume Not Met
Breath Volume		0.179	02:08	
Air Blank		0.000	02:08	
Control Test		0.078	02:09	RFI Detect
Air Blank		0.000	02:09	

Data Download Date/Time: 01/26/2016 09:36

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAYO,TREVOR,W
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: MAYO,TREVOR
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 12/29/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: RAJAH

DATE/TIME OF TEST: 01/23/2016 02:16:24

Diagnostics Check	OK	02:15	
Air Blank		02:15	RFI Detect
Air Blank	0.000	02:16	

Data Download Date/Time: 01/26/2016 09:36

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAYO,TREVOR,W
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: MAYO,TREVOR
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 12/29/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: RAJAH

DATE/TIME OF TEST: 01/23/2016 02:21:45

Diagnostics Check		OK	02:18	
Air Blank		0.000	02:18	
Control Test		0.081	02:18	
Air Blank		0.000	02:19	
Subject Sample	# 1		02:20	Subject Test Refused
Breath Volume		0.000	02:20	
Air Blank		0.000	02:20	
Control Test		0.081	02:21	
Air Blank		0.000	02:21	
Diagnostics Check		OK	02:21	

Data Download Date/Time: 01/26/2016 09:36

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUDSON,JOSHUA,K
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: HUDSON,JOSHUA
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 12/29/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:21

SUBJECT'S LAST NAME: BRIGGS

DATE/TIME OF TEST: 01/23/2016 03:58:24

Diagnostics Check		OK	03:50
Air Blank		0.000	03:50
Control Test		0.081	03:51
Air Blank		0.000	03:51
Subject Sample	# 1	0.245	03:53
Breath Volume		3.125	03:53
Air Blank		0.000	03:54
Air Blank		0.000	03:55
Subject Sample	# 2	0.238	03:56
Breath Volume		2.613	03:56
Air Blank		0.000	03:57
Control Test		0.080	03:57
Air Blank		0.000	03:58
Diagnostics Check		OK	03:58

Data Download Date/Time: 01/26/2016 09:36

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUDSON,JOSHUA,K
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: LOWRY,BRYAN
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 12/29/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:37

SUBJECT'S LAST NAME: KEWA

DATE/TIME OF TEST: 01/23/2016 22:12:34

Diagnostics Check		OK	22:04
Air Blank		0.000	22:04
Control Test		0.081	22:05
Air Blank		0.000	22:05
Subject Sample	# 1	0.208	22:06
Breath Volume		1.171	22:06
Air Blank		0.000	22:07
Air Blank		0.000	22:09
Subject Sample	# 2	0.203	22:10
Breath Volume		2.531	22:10
Air Blank		0.000	22:11
Control Test		0.081	22:11
Air Blank		0.000	22:12
Diagnostics Check		OK	22:12

Data Download Date/Time: 01/26/2016 09:36

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHORETTE,BRIAN,E
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: SHORETTE,BRIAN
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 12/29/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: PRICHARD

DATE/TIME OF TEST: 01/24/2016 00:42:22

Diagnostics Check		OK	00:35
Air Blank		0.000	00:36
Control Test		0.082	00:36
Air Blank		0.000	00:36
Subject Sample	# 1	0.086	00:37
Breath Volume		1.664	00:37
Air Blank		0.000	00:38
Air Blank		0.000	00:40
Subject Sample	# 2	0.085	00:40
Breath Volume		1.507	00:40
Air Blank		0.000	00:41
Control Test		0.079	00:41
Air Blank		0.000	00:42
Diagnostics Check		OK	00:42

Data Download Date/Time: 01/26/2016 09:36

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUDSON,JOSHUA,K
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: HUDSON,JOSHUA
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 12/29/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: MAYER

DATE/TIME OF TEST: 01/24/2016 01:38:43

Diagnostics Check		OK	01:30
Air Blank		0.000	01:31
Control Test		0.081	01:31
Air Blank		0.000	01:32
Subject Sample	# 1	0.242	01:33
Breath Volume		2.308	01:33
Air Blank		0.000	01:34
Air Blank		0.000	01:36
Subject Sample	# 2	0.237	01:37
Breath Volume		1.128	01:37
Air Blank		0.000	01:37
Control Test		0.080	01:38
Air Blank		0.000	01:38
Diagnostics Check		OK	01:38

Data Download Date/Time: 01/26/2016 09:36

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAYO,TREVOR,W
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: MAYO,TREVOR
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 12/29/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:47

SUBJECT'S LAST NAME: GILLEY

DATE/TIME OF TEST: 01/24/2016 04:24:33

Diagnostics Check		OK	04:08	
Air Blank		0.000	04:08	
Control Test		0.082	04:08	
Air Blank		0.000	04:09	
Subject Sample	# 1	0.143	04:12	Volume Not Met
Breath Volume		0.996	04:12	
Air Blank		0.000	04:13	
Air Blank		0.000	04:15	
Subject Sample	# 2	0.139	04:17	
Breath Volume		1.570	04:17	
Air Blank		0.000	04:17	
Air Blank		0.000	04:19	
Subject Sample	# 3	0.116	04:22	Volume Not Met
Breath Volume		0.812	04:22	
Air Blank		0.000	04:23	
Control Test		0.079	04:23	
Air Blank		0.000	04:24	
Diagnostics Check		OK	04:24	

Data Download Date/Time: 01/26/2016 09:36

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROOP,CHAD,E
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: ROOP,CHAD
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 01/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:44

SUBJECT'S LAST NAME: BRENNAN

DATE/TIME OF TEST: 01/28/2016 02:08:55

Diagnostics Check		OK	02:06	
Air Blank		0.000	02:06	
Control Test		0.081	02:06	
Air Blank		0.000	02:07	
Subject Sample	# 1		02:07	Subject Test Refused
Breath Volume		0.000	02:07	
Air Blank		0.000	02:07	
Control Test		0.081	02:08	
Air Blank		0.000	02:08	
Diagnostics Check		OK	02:08	

Data Download Date/Time: 02/23/2016 08:30

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SALON,DENNIS,M
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: SALON,DENNIS
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 01/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 15:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:36

SUBJECT'S LAST NAME: BRAKE

DATE/TIME OF TEST: 01/28/2016 16:03:40

Diagnostics Check		OK	15:57
Air Blank		0.000	15:57
Control Test		0.081	15:57
Air Blank		0.000	15:58
Subject Sample	# 1	0.000	15:59
Breath Volume		3.570	15:59
Air Blank		0.000	15:59
Air Blank		0.000	16:01
Subject Sample	# 2	0.000	16:02
Breath Volume		2.972	16:02
Air Blank		0.000	16:02
Control Test		0.079	16:03
Air Blank		0.000	16:03
Diagnostics Check		OK	16:03

Data Download Date/Time: 02/23/2016 08:30

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUDSON,JOSHUA,K
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: HUDSON,JOSHUA
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 01/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:08

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 01/29/2016 04:49:11

Diagnostics Check		OK	04:37	
Air Blank		0.000	04:38	
Control Test		0.080	04:38	
Air Blank		0.000	04:39	
Subject Sample	# 1		04:42	No Sample Provided
Breath Volume		0.000	04:42	
Air Blank		0.000	04:42	
Air Blank		0.000	04:44	
Subject Sample	# 2		04:47	No Sample Provided
Breath Volume		0.000	04:47	
Air Blank		0.000	04:48	
Control Test		0.080	04:48	
Air Blank		0.000	04:49	
Diagnostics Check		OK	04:49	

Data Download Date/Time: 02/23/2016 08:30

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEEKS,JEREMIAH,S
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: MEEKS,JEREMIAH
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 01/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:58

SUBJECT'S LAST NAME: DOMSCHKE

DATE/TIME OF TEST: 01/31/2016 00:24:56

Diagnostics Check		OK	00:18
Air Blank		0.000	00:19
Control Test		0.081	00:19
Air Blank		0.000	00:19
Subject Sample	# 1	0.000	00:20
Breath Volume		1.507	00:20
Air Blank		0.000	00:21
Air Blank		0.000	00:23
Subject Sample	# 2	0.000	00:23
Breath Volume		3.671	00:23
Air Blank		0.000	00:23
Control Test		0.079	00:24
Air Blank		0.000	00:24
Diagnostics Check		OK	00:24

Data Download Date/Time: 02/23/2016 08:30

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHORETTE,BRIAN,E
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: SHORETTE,BRIAN
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 01/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:44

SUBJECT'S LAST NAME: MORRIS

DATE/TIME OF TEST: 01/31/2016 01:10:34

Diagnostics Check		OK	01:04
Air Blank		0.000	01:04
Control Test		0.080	01:05
Air Blank		0.000	01:05
Subject Sample	# 1	0.114	01:06
Breath Volume		2.437	01:06
Air Blank		0.000	01:06
Air Blank		0.000	01:08
Subject Sample	# 2	0.113	01:08
Breath Volume		1.472	01:08
Air Blank		0.000	01:09
Control Test		0.079	01:09
Air Blank		0.000	01:10
Diagnostics Check		OK	01:10

Data Download Date/Time: 02/23/2016 08:30

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POOL,HAROLD,M
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: COTE,JACOB
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 01/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: GIBBS

DATE/TIME OF TEST: 01/31/2016 01:58:13

Diagnostics Check		OK	01:50
Air Blank		0.000	01:51
Control Test		0.080	01:51
Air Blank		0.000	01:52
Subject Sample	# 1	0.258	01:53
Breath Volume		2.367	01:53
Air Blank		0.000	01:53
Air Blank		0.000	01:55
Subject Sample	# 2	0.238	01:56
Breath Volume		2.070	01:56
Air Blank		0.000	01:57
Control Test		0.081	01:57
Air Blank		0.000	01:58
Diagnostics Check		OK	01:58

Data Download Date/Time: 02/23/2016 08:30

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POOL,HAROLD,M
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: ALVAREZ,ERIC
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 01/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: JUNTUNEN

DATE/TIME OF TEST: 01/31/2016 04:17:29

Diagnostics Check		OK	04:10
Air Blank		0.000	04:10
Control Test		0.081	04:11
Air Blank		0.000	04:11
Subject Sample	# 1	0.196	04:12
Breath Volume		1.171	04:12
Air Blank		0.000	04:13
Air Blank		0.000	04:15
Subject Sample	# 2	0.200	04:15
Breath Volume		1.234	04:15
Air Blank		0.000	04:16
Control Test		0.080	04:16
Air Blank		0.000	04:17
Diagnostics Check		OK	04:17

Data Download Date/Time: 02/23/2016 08:30

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NEWTON,JAMES,H
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: GREEN,HARRY
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 01/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 05:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:40

SUBJECT'S LAST NAME: COFFEY

DATE/TIME OF TEST: 01/31/2016 06:05:03

Diagnostics Check		OK	06:01	
Air Blank		0.000	06:02	
Control Test		0.080	06:02	
Air Blank		0.000	06:03	
Subject Sample	# 1		06:03	Subject Test Refused
Breath Volume		0.000	06:03	
Air Blank		0.000	06:04	
Control Test		0.081	06:04	
Air Blank		0.000	06:04	
Diagnostics Check		OK	06:05	

Data Download Date/Time: 02/23/2016 08:30

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEEKS,JEREMIAH,S
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: MAYO,NICHOLAS
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 01/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 11:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 11:28

SUBJECT'S LAST NAME: WARDEN

DATE/TIME OF TEST: 01/31/2016 11:56:42

Diagnostics Check		OK	11:48
Air Blank		0.000	11:48
Control Test		0.080	11:49
Air Blank		0.000	11:49
Subject Sample	# 1	0.253	11:52
Breath Volume		1.558	11:52
Air Blank		0.000	11:52
Air Blank		0.000	11:54
Subject Sample	# 2	0.259	11:55
Breath Volume		1.910	11:55
Air Blank		0.000	11:55
Control Test		0.078	11:56
Air Blank		0.000	11:56
Diagnostics Check		OK	11:56

Data Download Date/Time: 02/23/2016 08:30

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREENE,ROBERT,R
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: HORNE,CHASE
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 01/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:47

SUBJECT'S LAST NAME: MICKOW

DATE/TIME OF TEST: 01/31/2016 23:12:42

Diagnostics Check		OK	23:07	
Air Blank		0.000	23:08	
Control Test		0.080	23:08	
Air Blank		0.000	23:09	
Subject Sample	# 1		23:11	Subject Test Refused
Breath Volume		0.000	23:11	
Air Blank		0.000	23:11	
Control Test		0.080	23:12	
Air Blank		0.000	23:12	
Diagnostics Check		OK	23:12	

Data Download Date/Time: 02/23/2016 08:30

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,JASIEL,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: TARGIA,FRANK
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:21

SUBJECT'S LAST NAME: VIVAS

DATE/TIME OF TEST: 01/01/2016 22:52:42

Diagnostics Check		OK	22:42
Air Blank		0.000	22:43
Control Test		0.078	22:43
Air Blank		0.000	22:44
Subject Sample	# 1	0.219	22:46
Breath Volume		2.972	22:46
Air Blank		0.000	22:47
Air Blank		0.000	22:49
Subject Sample	# 2	0.211	22:51
Breath Volume		2.535	22:51
Air Blank		0.000	22:51
Control Test		0.078	22:52
Air Blank		0.000	22:52
Diagnostics Check		OK	22:52

Data Download Date/Time: 01/28/2016 15:18

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHESTER,CHARLES,W
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: CHESTER,CHARLES
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: RAPHAEL

DATE/TIME OF TEST: 01/03/2016 04:30:34

Diagnostics Check		OK	04:26	
Air Blank		0.000	04:26	
Control Test		0.077	04:27	
Air Blank		0.000	04:27	
Subject Sample	# 1		04:28	Subject Test Refused
Breath Volume		0.000	04:28	
Air Blank		0.000	04:29	
Control Test		0.079	04:29	
Air Blank		0.000	04:30	
Diagnostics Check		OK	04:30	

Data Download Date/Time: 01/28/2016 15:18

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLANCO,ROGER,J
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: AVILA,EDDIE
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: AMAYA

DATE/TIME OF TEST: 01/08/2016 00:16:41

Diagnostics Check		OK	00:13	
Air Blank		0.000	00:14	
Control Test		0.078	00:14	
Air Blank		0.000	00:14	
Subject Sample	# 1		00:15	Subject Test Refused
Breath Volume		0.000	00:15	
Air Blank		0.000	00:15	
Control Test		0.078	00:16	
Air Blank		0.000	00:16	
Diagnostics Check		OK	00:16	

Data Download Date/Time: 01/28/2016 15:18

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAZQUEZ,JONATHAN,L
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: MUINA,RAFAEL
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:43

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 01/09/2016 02:14:47

Diagnostics Check		OK	02:08
Air Blank		0.000	02:09
Control Test		0.078	02:09
Air Blank		0.000	02:09
Subject Sample	# 1	0.052	02:10
Breath Volume		2.187	02:10
Air Blank		0.000	02:11
Air Blank		0.000	02:12
Subject Sample	# 2	0.050	02:13
Breath Volume		2.839	02:13
Air Blank		0.000	02:13
Control Test		0.078	02:14
Air Blank		0.000	02:14
Diagnostics Check		OK	02:14

Data Download Date/Time: 01/28/2016 15:18

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TARGIA,FRANK,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: PALOMINO,JULIO
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: FERNANDEZ-MORIS

DATE/TIME OF TEST: 01/09/2016 03:20:24

Diagnostics Check		OK	03:11
Air Blank		0.000	03:12
Control Test		0.078	03:12
Air Blank		0.000	03:13
Subject Sample	# 1	0.131	03:15
Breath Volume		1.183	03:15
Air Blank		0.000	03:16
Air Blank		0.000	03:18
Subject Sample	# 2	0.130	03:18
Breath Volume		1.566	03:18
Air Blank		0.000	03:19
Control Test		0.075	03:19
Air Blank		0.000	03:20
Diagnostics Check		OK	03:20

Data Download Date/Time: 01/28/2016 15:18

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALOMINO,JULIO,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: TARGIA,FRANK
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:48

SUBJECT'S LAST NAME: CLIMACO

DATE/TIME OF TEST: 01/09/2016 03:41:16

Diagnostics Check		OK	03:28
Air Blank		0.000	03:28
Control Test		0.078	03:29
Air Blank		0.000	03:29
Subject Sample	# 1	0.213	03:31
Breath Volume		1.425	03:31
Air Blank		0.000	03:31
Air Blank		0.000	03:33
Subject Sample	# 2	0.191	03:35
Breath Volume		1.140	03:35
Air Blank		0.000	03:35
Air Blank		0.000	03:37
Subject Sample	# 3	0.190	03:39
Breath Volume		1.179	03:39
Air Blank		0.000	03:40
Control Test		0.076	03:40
Air Blank		0.000	03:41
Diagnostics Check		OK	03:41

Data Download Date/Time: 01/28/2016 15:18

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIL,JORGE,Y
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: GIL,JORGE
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:15

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 01/10/2016 06:43:02

Diagnostics Check		OK	06:38	
Air Blank		0.000	06:39	
Control Test		0.078	06:39	
Air Blank		0.000	06:39	
Subject Sample	# 1		06:41	Subject Test Refused
Breath Volume		0.000	06:41	
Air Blank		0.000	06:42	
Control Test		0.079	06:42	
Air Blank		0.000	06:42	
Diagnostics Check		OK	06:43	

Data Download Date/Time: 01/28/2016 15:18

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MATHIS,ANDRE,D
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: YEARBY,DASHON
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 09:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:36

SUBJECT'S LAST NAME: CASO OTERO

DATE/TIME OF TEST: 01/10/2016 10:13:11

Diagnostics Check		OK	10:01	
Air Blank		0.000	10:01	
Control Test		0.077	10:02	
Air Blank		0.000	10:02	
Subject Sample	# 1	0.153	10:05	Volume Not Met
Breath Volume		0.937	10:05	
Air Blank		0.000	10:06	
Air Blank		0.000	10:08	
Subject Sample	# 2	0.167	10:08	
Breath Volume		2.640	10:08	
Air Blank		0.000	10:09	
Air Blank		0.000	10:11	
Subject Sample	# 3	0.170	10:11	
Breath Volume		3.269	10:11	
Air Blank		0.000	10:12	
Control Test		0.074	10:12	Control Outside Tolerance
Air Blank		0.000	10:13	

Data Download Date/Time: 01/28/2016 15:18

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASTRO,JUAN,C
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: HERNANDEZ,CAMILO
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: RAMOS

DATE/TIME OF TEST: 01/10/2016 22:31:59

Diagnostics Check		OK	22:29	
Air Blank		0.000	22:29	
Control Test		0.078	22:29	
Air Blank		0.000	22:30	
Subject Sample	# 1		22:30	Subject Test Refused
Breath Volume		0.000	22:30	
Air Blank		0.000	22:31	
Control Test		0.078	22:31	
Air Blank		0.000	22:31	
Diagnostics Check		OK	22:31	

Data Download Date/Time: 01/28/2016 15:18

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GROSSMAN,JAY,J
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: GROSSMAN,JAY
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 18:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:10

SUBJECT'S LAST NAME: CRAWFORD

DATE/TIME OF TEST: 01/12/2016 19:34:36

Diagnostics Check		OK	19:31	
Air Blank		0.000	19:32	
Control Test		0.078	19:32	
Air Blank		0.000	19:33	
Subject Sample	# 1		19:33	Subject Test Refused
Breath Volume		0.000	19:33	
Air Blank		0.000	19:33	
Control Test		0.078	19:33	
Air Blank		0.000	19:34	
Diagnostics Check		OK	19:34	

Data Download Date/Time: 01/28/2016 15:18

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAPAZ,TAMMY,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: AMORES,NESTOR
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 01/14/2016 02:18:39

Diagnostics Check		OK	02:15	
Air Blank		0.000	02:16	
Control Test		0.075	02:16	
Air Blank		0.000	02:16	
Subject Sample	# 1		02:17	Subject Test Refused
Breath Volume		0.000	02:17	
Air Blank		0.000	02:17	
Control Test		0.076	02:18	
Air Blank		0.000	02:18	
Diagnostics Check		OK	02:18	

Data Download Date/Time: 01/28/2016 15:18

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PUJADAS,CHRISTOPHER,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: DANIER,P
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: HAILEY

DATE/TIME OF TEST: 01/15/2016 04:40:22

Diagnostics Check		OK	04:31	
Air Blank		0.000	04:32	
Control Test		0.078	04:32	
Air Blank		0.000	04:33	
Subject Sample	# 1		04:36	No Sample Provided
Breath Volume		0.000	04:36	
Air Blank		0.000	04:36	
Air Blank		0.000	04:38	
Subject Sample	# 2		04:38	Subject Test Refused
Breath Volume		0.000	04:38	
Air Blank		0.000	04:39	
Control Test		0.077	04:39	
Air Blank		0.000	04:40	
Diagnostics Check		OK	04:40	

Data Download Date/Time: 01/28/2016 15:18

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PUJADAS,CHRISTOPHER,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: COVERT,SHAWN
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: MARIN

DATE/TIME OF TEST: 01/23/2016 04:37:21

Diagnostics Check		OK	04:27
Air Blank		0.000	04:28
Control Test		0.078	04:28
Air Blank		0.000	04:29
Subject Sample	# 1	0.177	04:31
Breath Volume		2.714	04:31
Air Blank		0.000	04:31
Air Blank		0.000	04:33
Subject Sample	# 2	0.177	04:35
Breath Volume		2.476	04:35
Air Blank		0.000	04:36
Control Test		0.075	04:36
Air Blank		0.000	04:37
Diagnostics Check		OK	04:37

Data Download Date/Time: 01/28/2016 15:18

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TADLOCK,JAYNIS,L
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: CALABRESE,JOHN
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:54

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 01/23/2016 22:22:36

Diagnostics Check		OK	22:15
Air Blank		0.000	22:15
Control Test		0.076	22:16
Air Blank		0.000	22:16
Subject Sample	# 1	0.155	22:18
Breath Volume		1.199	22:18
Air Blank		0.000	22:18
Air Blank		0.000	22:20
Subject Sample	# 2	0.154	22:20
Breath Volume		1.429	22:20
Air Blank		0.000	22:21
Control Test		0.076	22:21
Air Blank		0.000	22:22
Diagnostics Check		OK	22:22

Data Download Date/Time: 01/28/2016 15:18

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HEREDIA RUBIO,ROBERTO,A
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: HEREDIA RUBIO,ROBERTO
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1131508A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 13:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:36

SUBJECT'S LAST NAME: MOULIN

DATE/TIME OF TEST: 01/24/2016 15:06:41

Diagnostics Check		OK	15:00
Air Blank		0.000	15:00
Control Test		0.076	15:01
Air Blank		0.000	15:01
Subject Sample	# 1	0.288	15:02
Breath Volume		2.464	15:02
Air Blank		0.000	15:02
Air Blank		0.000	15:04
Subject Sample	# 2	0.295	15:05
Breath Volume		2.445	15:05
Air Blank		0.000	15:05
Control Test		0.077	15:06
Air Blank		0.000	15:06
Diagnostics Check		OK	15:06

Data Download Date/Time: 01/28/2016 15:18

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINEZ,DAYAN,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: MARTINEZ,DAYAN
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 14:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:10

SUBJECT'S LAST NAME: BECKER

DATE/TIME OF TEST: 01/25/2016 15:40:05

Diagnostics Check		OK	15:33
Air Blank		0.000	15:33
Control Test		0.079	15:33
Air Blank		0.000	15:34
Subject Sample	# 1	0.262	15:35
Breath Volume		2.562	15:35
Air Blank		0.000	15:35
Air Blank		0.000	15:37
Subject Sample	# 2	0.268	15:38
Breath Volume		2.785	15:38
Air Blank		0.000	15:39
Control Test		0.077	15:39
Air Blank		0.000	15:40
Diagnostics Check		OK	15:40

Data Download Date/Time: 01/28/2016 15:18

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUSEL,BRIAN,M
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: HUSEL,BRIAN
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:19

SUBJECT'S LAST NAME: ROHATYNSKY

DATE/TIME OF TEST: 01/26/2016 21:47:32

Diagnostics Check		OK	21:44	
Air Blank		0.000	21:45	
Control Test		0.078	21:45	
Air Blank		0.000	21:45	
Subject Sample	# 1		21:46	Subject Test Refused
Breath Volume		0.000	21:46	
Air Blank		0.000	21:46	
Control Test		0.078	21:46	
Air Blank		0.000	21:47	
Diagnostics Check		OK	21:47	

Data Download Date/Time: 01/28/2016 15:18

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUNTER-RUBIO,ASHLEY,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: COLLEO,JACKIE
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 15:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:28

SUBJECT'S LAST NAME: FALCON

DATE/TIME OF TEST: 01/27/2016 15:51:53

Diagnostics Check		OK	15:48	
Air Blank		0.000	15:49	
Control Test		0.078	15:49	
Air Blank		0.000	15:50	
Subject Sample	# 1		15:50	Subject Test Refused
Breath Volume		0.000	15:50	
Air Blank		0.000	15:50	
Control Test		0.079	15:51	
Air Blank		0.000	15:51	
Diagnostics Check		OK	15:51	

Data Download Date/Time: 01/28/2016 15:18

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PUJADAS,CHRIS,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: PUJADAS,CHRIS
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 01/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: RAMBO

DATE/TIME OF TEST: 01/30/2016 01:25:29

Diagnostics Check		OK	01:19
Air Blank		0.000	01:19
Control Test		0.077	01:20
Air Blank		0.000	01:20
Subject Sample	# 1	0.131	01:21
Breath Volume		2.078	01:21
Air Blank		0.000	01:21
Air Blank		0.000	01:23
Subject Sample	# 2	0.132	01:23
Breath Volume		2.601	01:23
Air Blank		0.000	01:24
Control Test		0.077	01:24
Air Blank		0.000	01:25
Diagnostics Check		OK	01:25

Data Download Date/Time: 02/25/2016 18:07

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOMINGO,ROLANDO,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: DOMINGO,ROLANDO
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: CHRISTIANSEN

DATE/TIME OF TEST: 01/23/2016 01:48:03

Diagnostics Check		OK	01:45	
Air Blank		0.000	01:45	
Control Test		0.079	01:45	
Air Blank		0.000	01:46	
Subject Sample	# 1		01:46	Subject Test Refused
Breath Volume		0.000	01:46	
Air Blank		0.000	01:47	
Control Test		0.079	01:47	
Air Blank		0.000	01:47	
Diagnostics Check		OK	01:48	

Data Download Date/Time: 01/28/2016 15:39

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006473

INSTRUMENT REGISTERED TO: PINECREST PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRUZ,EDISON,B
BREATH TEST OPERATOR AGENCY: Village of Pinecrest
ARRESTING OFFICER: CRUZ,EDISON
ARRESTING OFFICER AGENCY: Village of Pinecrest
LAST AGENCY INSPECTION: 12/09/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:57

SUBJECT'S LAST NAME: REY

DATE/TIME OF TEST: 01/03/2016 05:25:01

Diagnostics Check		OK	05:18
Air Blank		0.000	05:18
Control Test		0.078	05:19
Air Blank		0.000	05:19
Subject Sample	# 1	0.108	05:20
Breath Volume		2.816	05:20
Air Blank		0.000	05:20
Air Blank		0.000	05:22
Subject Sample	# 2	0.105	05:23
Breath Volume		2.964	05:23
Air Blank		0.000	05:24
Control Test		0.077	05:24
Air Blank		0.000	05:24
Diagnostics Check		OK	05:25

Data Download Date/Time: 01/21/2016 09:39

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006473

INSTRUMENT REGISTERED TO: PINECREST PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,JOSE,L
BREATH TEST OPERATOR AGENCY: Village of Pinecrest
ARRESTING OFFICER: ALVAREZ,MICHAEL
ARRESTING OFFICER AGENCY: Village of Pinecrest
LAST AGENCY INSPECTION: 12/09/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 07:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:10

SUBJECT'S LAST NAME: AVILA

DATE/TIME OF TEST: 01/06/2016 08:41:20

Diagnostics Check		OK	08:35	
Air Blank		0.000	08:35	
Control Test		0.078	08:35	
Air Blank		0.000	08:36	
Subject Sample	# 1	0.176	08:36	
Breath Volume		2.234	08:36	
Air Blank		0.000	08:37	
Air Blank		0.000	08:39	
Subject Sample	# 2	0.163	08:39	
Breath Volume		1.574	08:39	
Air Blank		0.000	08:40	
Control Test		0.074	08:40	Control Outside Tolerance
Air Blank		0.000	08:41	

Data Download Date/Time: 01/21/2016 09:39

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006473

INSTRUMENT REGISTERED TO: PINECREST PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,JOSE,L
BREATH TEST OPERATOR AGENCY: Village of Pinecrest
ARRESTING OFFICER: ALVAREZ,MICHAEL
ARRESTING OFFICER AGENCY: Village of Pinecrest
LAST AGENCY INSPECTION: 12/09/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 07:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:45

SUBJECT'S LAST NAME: AVILA

DATE/TIME OF TEST: 01/06/2016 09:14:45

Diagnostics Check		OK	09:08	
Air Blank		0.000	09:08	
Control Test		0.077	09:08	
Air Blank		0.000	09:09	
Subject Sample	# 1	0.171	09:10	
Breath Volume		2.187	09:10	
Air Blank		0.000	09:10	
Air Blank		0.000	09:12	
Subject Sample	# 2	0.174	09:12	
Breath Volume		2.851	09:12	
Air Blank		0.000	09:13	
Control Test		0.073	09:14	Control Outside Tolerance
Air Blank		0.000	09:14	

Data Download Date/Time: 01/21/2016 09:39

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006473

INSTRUMENT REGISTERED TO: PINECREST PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIAZ,ALEXANDER,
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: BLANCO,JUAN
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 01/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: BUI
ARREST TIME: 20:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:33

SUBJECT'S LAST NAME: MUNOZ SOSA

DATE/TIME OF TEST: 01/21/2016 21:46:09

Diagnostics Check		OK	21:39
Air Blank		0.000	21:39
Control Test		0.078	21:39
Air Blank		0.000	21:40
Subject Sample	# 1	0.120	21:41
Breath Volume		2.175	21:41
Air Blank		0.000	21:42
Air Blank		0.000	21:43
Subject Sample	# 2	0.117	21:44
Breath Volume		1.445	21:44
Air Blank		0.000	21:45
Control Test		0.077	21:45
Air Blank		0.000	21:46
Diagnostics Check		OK	21:46

Data Download Date/Time: 02/12/2016 09:35

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006473

INSTRUMENT REGISTERED TO: PINECREST PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRUZ,EDISON,B
BREATH TEST OPERATOR AGENCY: Village of Pinecrest
ARRESTING OFFICER: CRUZ,EDISON
ARRESTING OFFICER AGENCY: Village of Pinecrest
LAST AGENCY INSPECTION: 01/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:02

SUBJECT'S LAST NAME: MARE

DATE/TIME OF TEST: 01/23/2016 04:28:02

Diagnostics Check		OK	04:24	
Air Blank		0.000	04:24	
Control Test		0.078	04:25	
Air Blank		0.000	04:25	
Subject Sample	# 1		04:26	Subject Test Refused
Breath Volume		0.000	04:26	
Air Blank		0.000	04:27	
Control Test		0.079	04:27	
Air Blank		0.000	04:27	
Diagnostics Check		OK	04:28	

Data Download Date/Time: 02/12/2016 09:35

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006473

INSTRUMENT REGISTERED TO: PINECREST PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRUZ,EDISON,B
BREATH TEST OPERATOR AGENCY: Village of Pinecrest
ARRESTING OFFICER: CRUZ,EDISON
ARRESTING OFFICER AGENCY: Village of Pinecrest
LAST AGENCY INSPECTION: 01/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:14

SUBJECT'S LAST NAME: SARRIA

DATE/TIME OF TEST: 01/28/2016 03:53:59

Diagnostics Check		OK	03:51	
Air Blank		0.000	03:51	
Control Test		0.080	03:51	
Air Blank		0.000	03:52	
Subject Sample	# 1		03:52	Subject Test Refused
Breath Volume		0.000	03:52	
Air Blank		0.000	03:53	
Control Test		0.078	03:53	
Air Blank		0.000	03:53	
Diagnostics Check		OK	03:53	

Data Download Date/Time: 02/12/2016 09:35

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARNANDAN,INDROWTIE,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GREETHAM,CHANDLER
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 02:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: SANON

DATE/TIME OF TEST: 01/01/2016 03:53:47

Diagnostics Check		OK	03:40	
Air Blank		0.000	03:40	
Control Test		0.080	03:41	
Air Blank		0.000	03:41	
Subject Sample	# 1	0.183	03:44	Volume Not Met
Breath Volume		0.070	03:44	
Air Blank		0.000	03:45	
Air Blank		0.000	03:47	
Subject Sample	# 2	0.207	03:48	
Breath Volume		1.476	03:48	
Air Blank		0.000	03:49	
Air Blank		0.000	03:51	
Subject Sample	# 3	0.215	03:52	
Breath Volume		1.343	03:52	
Air Blank		0.000	03:52	
Control Test		0.078	03:53	
Air Blank		0.000	03:53	
Diagnostics Check		OK	03:53	

Data Download Date/Time: 01/29/2016 18:50

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARNANDAN,INDROWTIE,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SERPHOS,STEVEN
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 03:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:57

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 01/02/2016 04:31:14

Diagnostics Check		OK	04:23
Air Blank		0.000	04:24
Control Test		0.079	04:24
Air Blank		0.000	04:24
Subject Sample	# 1	0.080	04:25
Breath Volume		1.406	04:25
Air Blank		0.000	04:26
Air Blank		0.000	04:28
Subject Sample	# 2	0.081	04:29
Breath Volume		1.695	04:29
Air Blank		0.000	04:30
Control Test		0.079	04:30
Air Blank		0.000	04:31
Diagnostics Check		OK	04:31

Data Download Date/Time: 01/29/2016 18:50

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SAPP,JOSHUA
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 05:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:53

SUBJECT'S LAST NAME: KERR

DATE/TIME OF TEST: 01/02/2016 06:21:39

Diagnostics Check		OK	06:15
Air Blank		0.000	06:15
Control Test		0.080	06:16
Air Blank		0.000	06:16
Subject Sample	# 1	0.129	06:17
Breath Volume		2.746	06:17
Air Blank		0.000	06:17
Air Blank		0.000	06:19
Subject Sample	# 2	0.124	06:20
Breath Volume		2.015	06:20
Air Blank		0.000	06:20
Control Test		0.080	06:21
Air Blank		0.000	06:21
Diagnostics Check		OK	06:21

Data Download Date/Time: 01/29/2016 18:50

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: STRZALKOWSKI,SEAN
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 00:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:09

SUBJECT'S LAST NAME: CROSBY

DATE/TIME OF TEST: 01/03/2016 01:36:57

Diagnostics Check		OK	01:29
Air Blank		0.000	01:30
Control Test		0.080	01:30
Air Blank		0.000	01:31
Subject Sample	# 1	0.112	01:32
Breath Volume		3.054	01:32
Air Blank		0.000	01:32
Air Blank		0.000	01:34
Subject Sample	# 2	0.108	01:35
Breath Volume		3.781	01:35
Air Blank		0.000	01:36
Control Test		0.079	01:36
Air Blank		0.000	01:36
Diagnostics Check		OK	01:36

Data Download Date/Time: 01/29/2016 18:50

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: PRESUME,JEHUD
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: CHAVEZ

DATE/TIME OF TEST: 01/03/2016 03:11:48

Diagnostics Check		OK	03:05
Air Blank		0.000	03:05
Control Test		0.080	03:06
Air Blank		0.000	03:06
Subject Sample	# 1	0.041	03:07
Breath Volume		3.468	03:07
Air Blank		0.000	03:07
Air Blank		0.000	03:09
Subject Sample	# 2	0.040	03:10
Breath Volume		2.964	03:10
Air Blank		0.000	03:10
Control Test		0.080	03:11
Air Blank		0.000	03:11
Diagnostics Check		OK	03:11

Data Download Date/Time: 01/29/2016 18:50

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: MANNING,MARK
ARRESTING OFFICER AGENCY: Pembroke Pines PD
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 03:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:12

SUBJECT'S LAST NAME: PINTO

DATE/TIME OF TEST: 01/03/2016 04:50:07

Diagnostics Check		OK	04:43
Air Blank		0.000	04:43
Control Test		0.080	04:43
Air Blank		0.000	04:44
Subject Sample	# 1	0.182	04:44
Breath Volume		3.144	04:44
Air Blank		0.000	04:45
Air Blank		0.000	04:47
Subject Sample	# 2	0.172	04:48
Breath Volume		2.722	04:48
Air Blank		0.000	04:49
Control Test		0.079	04:49
Air Blank		0.000	04:50
Diagnostics Check		OK	04:50

Data Download Date/Time: 01/29/2016 18:50

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: WILLIAMS,SEAN
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 18:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:30

SUBJECT'S LAST NAME: MCCLOUD

DATE/TIME OF TEST: 01/04/2016 19:58:03

Diagnostics Check		OK	19:51
Air Blank		0.000	19:52
Control Test		0.081	19:52
Air Blank		0.000	19:53
Subject Sample	# 1	0.000	19:53
Breath Volume		2.449	19:53
Air Blank		0.000	19:54
Air Blank		0.000	19:56
Subject Sample	# 2	0.000	19:56
Breath Volume		3.046	19:56
Air Blank		0.000	19:57
Control Test		0.080	19:57
Air Blank		0.000	19:57
Diagnostics Check		OK	19:58

Data Download Date/Time: 01/29/2016 18:50

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: CAROTTI,MOISES
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 18:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:56

SUBJECT'S LAST NAME: PATRICK

DATE/TIME OF TEST: 01/06/2016 19:32:00

Diagnostics Check		OK	19:16	
Air Blank		0.000	19:17	
Control Test		0.079	19:17	
Air Blank		0.000	19:18	
Subject Sample	# 1	0.313	19:19	
Breath Volume		1.343	19:19	
Air Blank		0.000	19:20	
Air Blank		0.000	19:21	
Subject Sample	# 2		19:24	No Sample Provided
Breath Volume		0.000	19:24	
Air Blank		0.000	19:25	
Air Blank		0.000	19:27	
Subject Sample	# 3		19:30	No Sample Provided
Breath Volume		0.000	19:30	
Air Blank		0.000	19:31	
Control Test		0.080	19:31	
Air Blank		0.000	19:31	
Diagnostics Check		OK	19:32	

Data Download Date/Time: 01/29/2016 18:50

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: CAROTTI,MOISES
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 18:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:35

SUBJECT'S LAST NAME: PATRICK

DATE/TIME OF TEST: 01/06/2016 20:02:12

Diagnostics Check		OK	19:55
Air Blank		0.000	19:56
Control Test		0.080	19:56
Air Blank		0.000	19:57
Subject Sample	# 1	0.288	19:57
Breath Volume		1.632	19:57
Air Blank		0.000	19:58
Air Blank		0.000	20:00
Subject Sample	# 2	0.285	20:00
Breath Volume		2.007	20:00
Air Blank		0.000	20:01
Control Test		0.080	20:01
Air Blank		0.000	20:02
Diagnostics Check		OK	20:02

Data Download Date/Time: 01/29/2016 18:50

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARNANDAN,INDROWTIE,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: CAROTTI,MOISES
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 01:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:09

SUBJECT'S LAST NAME: GINSBERG

DATE/TIME OF TEST: 01/07/2016 02:35:59

Diagnostics Check		OK	02:29
Air Blank		0.000	02:29
Control Test		0.080	02:30
Air Blank		0.000	02:30
Subject Sample	# 1	0.123	02:31
Breath Volume		1.761	02:31
Air Blank		0.000	02:31
Air Blank		0.000	02:33
Subject Sample	# 2	0.122	02:34
Breath Volume		2.078	02:34
Air Blank		0.000	02:35
Control Test		0.080	02:35
Air Blank		0.000	02:35
Diagnostics Check		OK	02:35

Data Download Date/Time: 01/29/2016 18:50

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: PERCIVAL,KAITLYN
ARRESTING OFFICER AGENCY: Pembroke Pines PD
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 02:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:22

SUBJECT'S LAST NAME: POLANCO

DATE/TIME OF TEST: 01/07/2016 03:54:19

Diagnostics Check		OK	03:47
Air Blank		0.000	03:47
Control Test		0.080	03:48
Air Blank		0.000	03:48
Subject Sample	# 1	0.175	03:49
Breath Volume		1.679	03:49
Air Blank		0.000	03:50
Air Blank		0.000	03:52
Subject Sample	# 2	0.172	03:52
Breath Volume		1.113	03:52
Air Blank		0.000	03:53
Control Test		0.079	03:53
Air Blank		0.000	03:54
Diagnostics Check		OK	03:54

Data Download Date/Time: 01/29/2016 18:50

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARNANDAN,INDROWTIE,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: LEE,TIMOTHY
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:46

SUBJECT'S LAST NAME: PHILLIPS

DATE/TIME OF TEST: 01/08/2016 02:12:56

Diagnostics Check		OK	02:06
Air Blank		0.000	02:06
Control Test		0.081	02:07
Air Blank		0.000	02:07
Subject Sample	# 1	0.226	02:08
Breath Volume		1.921	02:08
Air Blank		0.000	02:09
Air Blank		0.000	02:10
Subject Sample	# 2	0.244	02:11
Breath Volume		2.195	02:11
Air Blank		0.000	02:12
Control Test		0.079	02:12
Air Blank		0.000	02:12
Diagnostics Check		OK	02:12

Data Download Date/Time: 01/29/2016 18:50

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: LUNA,KRISTINA
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 23:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:58

SUBJECT'S LAST NAME: HOUSTON

DATE/TIME OF TEST: 01/09/2016 00:26:40

Diagnostics Check		OK	00:23	
Air Blank		0.000	00:24	
Control Test		0.081	00:24	
Air Blank		0.000	00:25	
Subject Sample	# 1		00:25	Subject Test Refused
Breath Volume		0.000	00:25	
Air Blank		0.000	00:25	
Control Test		0.081	00:26	
Air Blank		0.000	00:26	
Diagnostics Check		OK	00:26	

Data Download Date/Time: 01/29/2016 18:50

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: WIENEINGER,JOSH
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 18:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:22

SUBJECT'S LAST NAME: SOSA-VARAS

DATE/TIME OF TEST: 01/09/2016 19:53:45

Diagnostics Check		OK	19:45
Air Blank		0.000	19:46
Control Test		0.081	19:46
Air Blank		0.000	19:47
Subject Sample	# 1	0.140	19:48
Breath Volume		1.128	19:48
Air Blank		0.000	19:49
Air Blank		0.000	19:51
Subject Sample	# 2	0.139	19:52
Breath Volume		1.535	19:52
Air Blank		0.000	19:52
Control Test		0.080	19:53
Air Blank		0.000	19:53
Diagnostics Check		OK	19:53

Data Download Date/Time: 01/29/2016 18:50

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUN,TONY,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SHAW,DANIEL
ARRESTING OFFICER AGENCY: Margate PD
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 13:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:15

SUBJECT'S LAST NAME: ULYSSE

DATE/TIME OF TEST: 01/10/2016 14:59:56

Diagnostics Check		OK	14:47	
Air Blank		0.000	14:47	
Control Test		0.081	14:47	
Air Blank		0.000	14:48	
Subject Sample	# 1	0.144	14:51	Volume Not Met
Breath Volume		0.582	14:51	
Air Blank		0.000	14:52	
Air Blank		0.000	14:53	
Subject Sample	# 2	0.200	14:54	
Breath Volume		1.835	14:54	
Air Blank		0.000	14:55	
Air Blank		0.000	14:56	
Subject Sample	# 3	0.205	14:58	
Breath Volume		1.500	14:58	
Air Blank		0.000	14:58	
Control Test		0.080	14:59	
Air Blank		0.000	14:59	
Diagnostics Check		OK	14:59	

Data Download Date/Time: 01/29/2016 18:50

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: PALANT,JASON
ARRESTING OFFICER AGENCY: Pembroke Pines PD
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 00:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:12

SUBJECT'S LAST NAME: BERON CARDONA

DATE/TIME OF TEST: 01/14/2016 01:56:09

Diagnostics Check		OK	01:48
Air Blank		0.000	01:49
Control Test		0.080	01:49
Air Blank		0.000	01:50
Subject Sample	# 1	0.085	01:51
Breath Volume		1.886	01:51
Air Blank		0.000	01:51
Air Blank		0.000	01:53
Subject Sample	# 2	0.085	01:54
Breath Volume		1.671	01:54
Air Blank		0.000	01:55
Control Test		0.080	01:55
Air Blank		0.000	01:56
Diagnostics Check		OK	01:56

Data Download Date/Time: 01/29/2016 18:50

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ECKINGER,LANI,A
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: BLUM,JOHN
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 07:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:37

SUBJECT'S LAST NAME: BAPTISTE

DATE/TIME OF TEST: 01/15/2016 09:06:15

Diagnostics Check		OK	09:03	
Air Blank		0.000	09:03	
Control Test		0.080	09:04	
Air Blank		0.000	09:04	
Subject Sample	# 1		09:04	Subject Test Refused
Breath Volume		0.000	09:04	
Air Blank		0.000	09:05	
Control Test		0.080	09:05	
Air Blank		0.000	09:06	
Diagnostics Check		OK	09:06	

Data Download Date/Time: 01/29/2016 18:50

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ECKINGER,LANI,A
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: BLUM,JOHN
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 07:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:44

SUBJECT'S LAST NAME: BAPTISTE

DATE/TIME OF TEST: 01/15/2016 09:14:48

Diagnostics Check		OK	09:08
Air Blank		0.000	09:08
Control Test		0.080	09:09
Air Blank		0.000	09:09
Subject Sample	# 1	0.162	09:10
Breath Volume		2.636	09:10
Air Blank		0.000	09:10
Air Blank		0.000	09:12
Subject Sample	# 2	0.155	09:13
Breath Volume		1.824	09:13
Air Blank		0.000	09:13
Control Test		0.081	09:14
Air Blank		0.000	09:14
Diagnostics Check		OK	09:14

Data Download Date/Time: 01/29/2016 18:50

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARNANDAN,INDROWTIE,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: RAUSCH,ROBERT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 05:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:02

SUBJECT'S LAST NAME: ALARCON

DATE/TIME OF TEST: 01/16/2016 06:29:05

Diagnostics Check		OK	06:22
Air Blank		0.000	06:22
Control Test		0.080	06:23
Air Blank		0.000	06:23
Subject Sample	# 1	0.205	06:24
Breath Volume		1.765	06:24
Air Blank		0.000	06:25
Air Blank		0.000	06:26
Subject Sample	# 2	0.210	06:27
Breath Volume		2.062	06:27
Air Blank		0.000	06:28
Control Test		0.079	06:28
Air Blank		0.000	06:29
Diagnostics Check		OK	06:29

Data Download Date/Time: 01/29/2016 18:50

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINA,MARMILLY,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SMITH,AARON
ARRESTING OFFICER AGENCY: Hallandale Beach PD
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 08:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:45

SUBJECT'S LAST NAME: LASSUS

DATE/TIME OF TEST: 01/16/2016 10:23:47

Diagnostics Check		OK	10:16
Air Blank		0.000	10:16
Control Test		0.080	10:16
Air Blank		0.000	10:17
Subject Sample	# 1	0.188	10:18
Breath Volume		1.277	10:18
Air Blank		0.000	10:19
Air Blank		0.000	10:21
Subject Sample	# 2	0.191	10:22
Breath Volume		1.449	10:22
Air Blank		0.000	10:22
Control Test		0.080	10:23
Air Blank		0.000	10:23
Diagnostics Check		OK	10:23

Data Download Date/Time: 01/29/2016 18:50

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUN,TONY,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: LEE,TIMOTHY
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 23:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:41

SUBJECT'S LAST NAME: PUERTA

DATE/TIME OF TEST: 01/17/2016 00:11:11

Diagnostics Check		OK	00:01
Air Blank		0.000	00:02
Control Test		0.078	00:02
Air Blank		0.000	00:03
Subject Sample	# 1	0.211	00:05
Breath Volume		1.304	00:05
Air Blank		0.000	00:06
Air Blank		0.000	00:07
Subject Sample	# 2	0.202	00:09
Breath Volume		1.148	00:09
Air Blank		0.000	00:10
Control Test		0.077	00:10
Air Blank		0.000	00:11
Diagnostics Check		OK	00:11

Data Download Date/Time: 01/29/2016 18:50

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINA,MARMILLY,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: POPICK,SCOTT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 04:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:15

SUBJECT'S LAST NAME: ZAPPIA

DATE/TIME OF TEST: 01/19/2016 05:45:20

Diagnostics Check		OK	05:38
Air Blank		0.000	05:39
Control Test		0.081	05:39
Air Blank		0.000	05:40
Subject Sample	# 1	0.140	05:40
Breath Volume		2.898	05:40
Air Blank		0.000	05:41
Air Blank		0.000	05:43
Subject Sample	# 2	0.134	05:43
Breath Volume		1.878	05:43
Air Blank		0.000	05:44
Control Test		0.081	05:44
Air Blank		0.000	05:45
Diagnostics Check		OK	05:45

Data Download Date/Time: 01/29/2016 18:50

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DICIENZO,FRANCA,R
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: MARTINEZ,YANSEL
ARRESTING OFFICER AGENCY: Hollywood PD
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 16:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:57

SUBJECT'S LAST NAME: FREEMAN

DATE/TIME OF TEST: 01/19/2016 17:31:22

Diagnostics Check		OK	17:24
Air Blank		0.000	17:25
Control Test		0.081	17:25
Air Blank		0.000	17:25
Subject Sample	# 1	0.173	17:26
Breath Volume		2.218	17:26
Air Blank		0.000	17:27
Air Blank		0.000	17:29
Subject Sample	# 2	0.165	17:29
Breath Volume		1.242	17:29
Air Blank		0.000	17:30
Control Test		0.081	17:30
Air Blank		0.000	17:31
Diagnostics Check		OK	17:31

Data Download Date/Time: 01/29/2016 18:50

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: HARRIS,DELLICA
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 21:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:48

SUBJECT'S LAST NAME: LANZA

DATE/TIME OF TEST: 01/19/2016 22:26:12

Diagnostics Check		OK	22:19
Air Blank		0.000	22:20
Control Test		0.081	22:20
Air Blank		0.000	22:20
Subject Sample	# 1	0.254	22:21
Breath Volume		1.382	22:21
Air Blank		0.000	22:22
Air Blank		0.000	22:23
Subject Sample	# 2	0.249	22:24
Breath Volume		1.398	22:24
Air Blank		0.000	22:25
Control Test		0.080	22:25
Air Blank		0.000	22:26
Diagnostics Check		OK	22:26

Data Download Date/Time: 01/29/2016 18:50

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: MAHMOUD,AYHAM
ARRESTING OFFICER AGENCY: Seminole Dept of LE
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 00:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: JOISEUS

DATE/TIME OF TEST: 01/20/2016 02:47:48

Diagnostics Check		OK	02:41
Air Blank		0.000	02:41
Control Test		0.081	02:42
Air Blank		0.000	02:42
Subject Sample	# 1	0.000	02:43
Breath Volume		2.679	02:43
Air Blank		0.000	02:43
Air Blank		0.000	02:45
Subject Sample	# 2	0.000	02:46
Breath Volume		2.687	02:46
Air Blank		0.000	02:46
Control Test		0.081	02:47
Air Blank		0.000	02:47
Diagnostics Check		OK	02:47

Data Download Date/Time: 01/29/2016 18:50

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: MEYERS,ERIC
ARRESTING OFFICER AGENCY: Sunrise PD
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:03

SUBJECT'S LAST NAME: FASULO

DATE/TIME OF TEST: 01/23/2016 01:43:05

Diagnostics Check		OK	01:29	
Air Blank		0.000	01:30	
Control Test		0.080	01:30	
Air Blank		0.000	01:31	
Subject Sample	# 1	0.242	01:32	No .020 agreement
Breath Volume		2.710	01:32	
Air Blank		0.000	01:33	
Air Blank		0.000	01:35	
Subject Sample	# 2	0.218	01:36	No .020 agreement
Breath Volume		1.839	01:36	
Air Blank		0.000	01:36	
Air Blank		0.000	01:38	
Subject Sample	# 3		01:41	No Sample Provided
Breath Volume		0.000	01:41	
Air Blank		0.000	01:42	
Control Test		0.079	01:42	
Air Blank		0.000	01:43	
Diagnostics Check		OK	01:43	

Data Download Date/Time: 01/29/2016 18:50

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ABRAMS,JESSICA
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 02:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:33

SUBJECT'S LAST NAME: IGLESIAS

DATE/TIME OF TEST: 01/23/2016 04:08:11

Diagnostics Check		OK	03:59
Air Blank		0.000	04:00
Control Test		0.080	04:00
Air Blank		0.000	04:01
Subject Sample	# 1	0.142	04:03
Breath Volume		2.539	04:03
Air Blank		0.000	04:03
Air Blank		0.000	04:05
Subject Sample	# 2	0.142	04:06
Breath Volume		2.144	04:06
Air Blank		0.000	04:07
Control Test		0.079	04:07
Air Blank		0.000	04:08
Diagnostics Check		OK	04:08

Data Download Date/Time: 01/29/2016 18:50

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARNANDAN,INDROWTIE,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: DERIET,RAFAEL
ARRESTING OFFICER AGENCY: Hollywood PD
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 05:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:12

SUBJECT'S LAST NAME: MASSEY

DATE/TIME OF TEST: 01/28/2016 06:44:43

Diagnostics Check		OK	06:33	
Air Blank		0.000	06:33	
Control Test		0.081	06:33	
Air Blank		0.000	06:34	
Subject Sample	# 1		06:37	No Sample Provided
Breath Volume		0.000	06:37	
Air Blank		0.000	06:38	
Air Blank		0.000	06:39	
Subject Sample	# 2		06:43	No Sample Provided
Breath Volume		0.000	06:43	
Air Blank		0.000	06:43	
Control Test		0.079	06:44	
Air Blank		0.000	06:44	
Diagnostics Check		OK	06:44	

Data Download Date/Time: 01/29/2016 18:50

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DONOHUE,LAWRENCE,E
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: DONOHUE,LAWRENCE
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 01/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 21:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: AUSSENHOFER

DATE/TIME OF TEST: 01/26/2016 23:17:39

Diagnostics Check		OK	23:11
Air Blank		0.000	23:11
Control Test		0.079	23:12
Air Blank		0.000	23:12
Subject Sample	# 1	0.000	23:13
Breath Volume		2.574	23:13
Air Blank		0.000	23:13
Air Blank		0.000	23:15
Subject Sample	# 2	0.000	23:16
Breath Volume		1.867	23:16
Air Blank		0.000	23:16
Control Test		0.075	23:17
Air Blank		0.000	23:17
Diagnostics Check		OK	23:17

Data Download Date/Time: 02/26/2016 10:03

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PETERSON,STEFANIE,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: PETERSON,STEFANIE
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 01/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 00:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: VASQUEZ

DATE/TIME OF TEST: 01/27/2016 02:42:24

Diagnostics Check		OK	02:35
Air Blank		0.000	02:36
Control Test		0.078	02:36
Air Blank		0.000	02:36
Subject Sample	# 1	0.110	02:37
Breath Volume		3.441	02:37
Air Blank		0.000	02:38
Air Blank		0.000	02:40
Subject Sample	# 2	0.104	02:40
Breath Volume		2.234	02:40
Air Blank		0.000	02:41
Control Test		0.078	02:41
Air Blank		0.000	02:42
Diagnostics Check		OK	02:42

Data Download Date/Time: 02/26/2016 10:03

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COROMINAS,BENNY,
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: COROMINAS,COROMINAS
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 01/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 13:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:30

SUBJECT'S LAST NAME: BURDEN

DATE/TIME OF TEST: 01/27/2016 13:57:33

Diagnostics Check		OK	13:51
Air Blank		0.000	13:52
Control Test		0.078	13:52
Air Blank		0.000	13:52
Subject Sample	# 1	0.000	13:53
Breath Volume		2.042	13:53
Air Blank		0.000	13:53
Air Blank		0.000	13:55
Subject Sample	# 2	0.000	13:56
Breath Volume		2.437	13:56
Air Blank		0.000	13:56
Control Test		0.075	13:56
Air Blank		0.000	13:57
Diagnostics Check		OK	13:57

Data Download Date/Time: 02/26/2016 10:03

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLE,DENNIS,A
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: COLE,DENNIS
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 01/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: DANIELS

DATE/TIME OF TEST: 01/28/2016 03:58:29

Diagnostics Check		OK	03:51
Air Blank		0.000	03:52
Control Test		0.078	03:52
Air Blank		0.000	03:53
Subject Sample	# 1	0.191	03:53
Breath Volume		2.789	03:53
Air Blank		0.000	03:54
Air Blank		0.000	03:56
Subject Sample	# 2	0.186	03:56
Breath Volume		1.292	03:56
Air Blank		0.000	03:57
Control Test		0.077	03:57
Air Blank		0.000	03:58
Diagnostics Check		OK	03:58

Data Download Date/Time: 02/26/2016 10:03

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 01/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 19:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: HABERLAND

DATE/TIME OF TEST: 01/29/2016 20:28:03

Diagnostics Check		OK	20:21
Air Blank		0.000	20:21
Control Test		0.079	20:21
Air Blank		0.000	20:22
Subject Sample	# 1	0.158	20:23
Breath Volume		2.937	20:23
Air Blank		0.000	20:24
Air Blank		0.000	20:25
Subject Sample	# 2	0.152	20:26
Breath Volume		2.882	20:26
Air Blank		0.000	20:27
Control Test		0.075	20:27
Air Blank		0.000	20:27
Diagnostics Check		OK	20:28

Data Download Date/Time: 02/26/2016 10:03

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: MEYER,ROBERT
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 01/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 03:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:41

SUBJECT'S LAST NAME: BOYLE

DATE/TIME OF TEST: 01/30/2016 05:18:46

Diagnostics Check		OK	05:11
Air Blank		0.000	05:11
Control Test		0.078	05:11
Air Blank		0.000	05:12
Subject Sample	# 1	0.242	05:13
Breath Volume		1.500	05:13
Air Blank		0.000	05:14
Air Blank		0.000	05:16
Subject Sample	# 2	0.259	05:17
Breath Volume		1.511	05:17
Air Blank		0.000	05:17
Control Test		0.077	05:18
Air Blank		0.000	05:18
Diagnostics Check		OK	05:18

Data Download Date/Time: 02/26/2016 10:03

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 01/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 00:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 01/31/2016 02:07:50

Diagnostics Check		OK	02:00	
Air Blank		0.000	02:00	
Control Test		0.080	02:01	
Air Blank		0.000	02:01	
Subject Sample	# 1	0.222	02:02	
Breath Volume		2.687	02:02	
Air Blank		0.000	02:02	
Air Blank		0.000	02:04	
Subject Sample	# 2		02:06	Subject Test Refused
Breath Volume		0.000	02:06	
Air Blank		0.000	02:06	
Control Test		0.077	02:07	
Air Blank		0.000	02:07	
Diagnostics Check		OK	02:07	

Data Download Date/Time: 02/26/2016 10:03

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.