

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000206

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAGNER,ROBERT,F
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: WAGNER,ROBERT
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 12/03/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG400406
DRY GAS EXPIRY DATE: 01/06/2016
VIOLATION CODE: DUI
ARREST TIME: 16:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:55

SUBJECT'S LAST NAME: PHILLIPS

DATE/TIME OF TEST: 01/01/2016 17:25:35

Diagnostics Check	OK	17:19
Air Blank	0.000	17:20
Control Test	0.079	17:20
Air Blank	0.000	17:21
Subject Sample # 1	0.000	17:21
Breath Volume	2.230	17:21
Air Blank	0.000	17:21
Air Blank	0.000	17:23
Subject Sample # 2	0.000	17:24
Breath Volume	2.183	17:24
Air Blank	0.000	17:24
Control Test	0.080	17:25
Air Blank	0.000	17:25
Diagnostics Check	OK	17:25

Data Download Date/Time: 01/22/2016 07:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000206

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASE, THOMAS, D
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: JENKINS, JOSEPH
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 12/03/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG425202
DRY GAS EXPIRY DATE: 09/09/2016
VIOLATION CODE: DUI
ARREST TIME: 23:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:05

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 01/15/2016 00:36:44

Diagnostics Check	OK	00:20	
Air Blank	0.000	00:21	
Control Test	0.079	00:21	
Air Blank	0.000	00:22	
Subject Sample	# 1 0.225	00:23	
Breath Volume	1.148	00:23	
Air Blank	0.000	00:24	
Air Blank	0.000	00:26	
Subject Sample	# 2 0.195	00:29	Volume Not Met
Breath Volume	0.746	00:29	
Air Blank	0.000	00:30	
Air Blank	0.000	00:32	
Subject Sample	# 3 0.176	00:35	Volume Not Met
Breath Volume	0.453	00:35	
Air Blank	0.000	00:35	
Control Test	0.079	00:36	
Air Blank	0.000	00:36	
Diagnostics Check	OK	00:36	

Data Download Date/Time: 01/22/2016 07:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000206

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: CASE,THOMAS,D

BREATH TEST OPERATOR AGENCY: Brevard County SO

ARRESTING OFFICER: LOVELL,DAVID

ARRESTING OFFICER AGENCY: Brevard County SO

LAST AGENCY INSPECTION: 12/03/2015

Date or Time Changed: No

DRY GAS STANDARD LOT #: AG425202

DRY GAS EXPIRY DATE: 09/09/2015

VIOLATION CODE: PROB/PAROL

ARREST TIME: 02:14

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:06

SUBJECT'S LAST NAME: STOREY

DATE/TIME OF TEST: 01/15/2016 02:40:40

Diagnostics Check	OK	02:28
Air Blank	0.000	02:28
Control Test	0.078	02:28
Air Blank	0.000	02:29
Subject Sample # 1	0.198	02:31
Breath Volume	1.144	02:31
Air Blank	0.000	02:32
Air Blank	0.000	02:34
Subject Sample # 2	0.223	02:34
Breath Volume	1.652	02:34
Air Blank	0.000	02:35
Air Blank	0.000	02:37
Subject Sample # 3	0.214	02:39
Breath Volume	1.976	02:39
Air Blank	0.000	02:39
Control Test	0.079	02:40
Air Blank	0.000	02:40
Diagnostics Check	OK	02:40

Data Download Date/Time: 01/22/2016 07:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000206

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAGNER,ROBERT,F
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: WAGNER,ROBERT
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 12/03/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG425202
DRY GAS EXPIRY DATE: 09/09/2016
VIOLATION CODE: COURT ORD
ARREST TIME: 08:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:15

SUBJECT'S LAST NAME: PHILLIPS

DATE/TIME OF TEST: 01/21/2016 08:43:54

Diagnostics Check	OK	08:38
Air Blank	0.000	08:38
Control Test	0.079	08:38
Air Blank	0.000	08:39
Subject Sample # 1	0.000	08:39
Breath Volume	2.695	08:39
Air Blank	0.000	08:40
Air Blank	0.000	08:42
Subject Sample # 2	0.000	08:42
Breath Volume	2.152	08:42
Air Blank	0.000	08:43
Control Test	0.078	08:43
Air Blank	0.000	08:43
Diagnostics Check	OK	08:43

Data Download Date/Time: 01/22/2016 07:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000206

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: GARRITY, THOMAS, A

BREATH TEST OPERATOR AGENCY: Brevard County SO

ARRESTING OFFICER: GLEASON, WILLIAM

ARRESTING OFFICER AGENCY: Brevard County SO

LAST AGENCY INSPECTION: 12/03/2015

Date or Time Changed: No

DRY GAS STANDARD LOT #: AG425202

DRY GAS EXPIRY DATE: 09/09/2016

VIOLATION CODE: DUI

ARREST TIME: 21:10

GENDER: Male

OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: SNYDER

DATE/TIME OF TEST: 01/21/2016 22:07:57

Diagnostics Check	OK	22:01
Air Blank	0.000	22:01
Control Test	0.078	22:01
Air Blank	0.000	22:02
Subject Sample # 1	0.184	22:03
Breath Volume	3.000	22:03
Air Blank	0.000	22:03
Air Blank	0.000	22:05
Subject Sample # 2	0.194	22:06
Breath Volume	3.386	22:06
Air Blank	0.000	22:07
Control Test	0.079	22:07
Air Blank	0.000	22:07
Diagnostics Check	OK	22:07

Data Download Date/Time: 01/22/2016 07:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRICKER,JEFFREY,S
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: FRICKER,JEFFREY
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 12/24/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: HOLDER

DATE/TIME OF TEST: 01/01/2016 03:04:38

Diagnostics Check	OK	02:56	
Air Blank	0.000	02:56	
Control Test	0.078	02:56	
Air Blank	0.000	02:57	
Subject Sample	# 1	03:00	No Sample Provided
Breath Volume	0.000	03:00	
Air Blank	0.000	03:00	
Air Blank	0.000	03:02	
Subject Sample	# 2	03:03	Subject Test Refused
Breath Volume	0.000	03:03	
Air Blank	0.000	03:03	
Control Test	0.077	03:04	
Air Blank	0.000	03:04	
Diagnostics Check	OK	03:04	

Data Download Date/Time: 01/22/2016 10:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000738

INSTRUMENT REGISTERED TO: FT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAIN,DANIEL,A
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: CAIN,DANIEL
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A1
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 19:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:40

SUBJECT'S LAST NAME: CULIC

DATE/TIME OF TEST: 01/03/2016 20:05:02

Diagnostics Check	OK	20:01	
Air Blank	0.000	20:01	
Control Test	0.081	20:01	
Air Blank	0.000	20:02	
Subject Sample	# 1	20:03	Subject Test Refused
Breath Volume	0.000	20:03	
Air Blank	0.000	20:04	
Control Test	0.080	20:04	
Air Blank	0.000	20:04	
Diagnostics Check	OK	20:05	

Data Download Date/Time: 01/31/2016 14:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000738

INSTRUMENT REGISTERED TO: FT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PETTIS,CHARLES,A
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: PETTIS,CHARLES
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A1
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:46

SUBJECT'S LAST NAME: HUYNH

DATE/TIME OF TEST: 01/03/2016 22:20:35

Diagnostics Check	OK	22:12
Air Blank	0.000	22:13
Control Test	0.080	22:13
Air Blank	0.000	22:13
Subject Sample # 1	0.134	22:15
Breath Volume	1.687	22:15
Air Blank	0.000	22:16
Air Blank	0.000	22:17
Subject Sample # 2	0.127	22:19
Breath Volume	1.703	22:19
Air Blank	0.000	22:19
Control Test	0.078	22:19
Air Blank	0.000	22:20
Diagnostics Check	OK	22:20

Data Download Date/Time: 01/31/2016 14:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000738

INSTRUMENT REGISTERED TO: FT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAILOR,TRAVIS,C
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: SAILOR,TRAVIS
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A1
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:24

SUBJECT'S LAST NAME: STARR II

DATE/TIME OF TEST: 01/12/2016 00:01:28

Diagnostics Check	OK	23:58	
Air Blank	0.000	23:59	
Control Test	0.081	23:59	
Air Blank	0.000	23:59	
Subject Sample	# 1	00:00	Subject Test Refused
Breath Volume	0.000	00:00	
Air Blank	0.000	00:00	
Control Test	0.080	00:00	
Air Blank	0.000	00:01	
Diagnostics Check	OK	00:01	

Data Download Date/Time: 01/31/2016 14:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000738

INSTRUMENT REGISTERED TO: FT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRAWFORD,BENJAMIN,D
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: CRAWFORD,BENJAMIN
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A1
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 06:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:52

SUBJECT'S LAST NAME: SCOTT

DATE/TIME OF TEST: 01/16/2016 07:22:22

Diagnostics Check	OK	07:15
Air Blank	0.000	07:16
Control Test	0.080	07:16
Air Blank	0.000	07:17
Subject Sample # 1	0.154	07:17
Breath Volume	1.433	07:17
Air Blank	0.000	07:18
Air Blank	0.000	07:20
Subject Sample # 2	0.159	07:20
Breath Volume	2.132	07:20
Air Blank	0.000	07:21
Control Test	0.079	07:21
Air Blank	0.000	07:22
Diagnostics Check	OK	07:22

Data Download Date/Time: 01/31/2016 14:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000738

INSTRUMENT REGISTERED TO: FT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRUJILLO,ANTHONY,R
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: TRUJILLO,ANTHONY
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A1
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 18:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:04

SUBJECT'S LAST NAME: DAWS

DATE/TIME OF TEST: 01/16/2016 19:27:49

Diagnostics Check	OK	19:25	
Air Blank	0.000	19:25	
Control Test	0.080	19:25	
Air Blank	0.000	19:26	
Subject Sample	# 1	19:26	Subject Test Refused
Breath Volume	0.000	19:26	
Air Blank	0.000	19:26	
Control Test	0.079	19:27	
Air Blank	0.000	19:27	
Diagnostics Check	OK	19:27	

Data Download Date/Time: 01/31/2016 14:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000738

INSTRUMENT REGISTERED TO: FT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESNEUL,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: ESNEUL,MICHAEL
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A1
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:29

SUBJECT'S LAST NAME: FABIAN

DATE/TIME OF TEST: 01/17/2016 05:01:49

Diagnostics Check	OK	04:50	
Air Blank	0.000	04:50	
Control Test	0.080	04:51	
Air Blank	0.000	04:51	
Subject Sample	# 1	04:54	No Sample Provided
Breath Volume	0.000	04:54	
Air Blank	0.000	04:55	
Air Blank	0.000	04:57	
Subject Sample	# 2	05:00	No Sample Provided
Breath Volume	0.000	05:00	
Air Blank	0.000	05:00	
Control Test	0.081	05:01	
Air Blank	0.000	05:01	
Diagnostics Check	OK	05:01	

Data Download Date/Time: 01/31/2016 14:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000738

INSTRUMENT REGISTERED TO: FT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESNEUL,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: ESNEUL,MICHAEL
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A1
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:29

SUBJECT'S LAST NAME: FABIAN

DATE/TIME OF TEST: 01/17/2016 05:06:31

Diagnostics Check	OK	05:03	
Air Blank	0.000	05:04	
Control Test	0.080	05:04	
Air Blank	0.000	05:04	
Subject Sample	# 1	05:05	Subject Test Refused
Breath Volume	0.000	05:05	
Air Blank	0.000	05:05	
Control Test	0.080	05:05	
Air Blank	0.000	05:06	
Diagnostics Check	OK	05:06	

Data Download Date/Time: 01/31/2016 14:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000738

INSTRUMENT REGISTERED TO: FT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPEAKMAN,ROBERT,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: SPEAKMAN,ROBERT
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A1
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: SOLYMOSI

DATE/TIME OF TEST: 01/23/2016 03:59:05

Diagnostics Check	OK	03:47	
Air Blank	0.000	03:47	
Control Test	0.081	03:48	
Air Blank	0.000	03:48	
Subject Sample	# 1 0.122	03:51	Volume Not Met
Breath Volume	0.207	03:51	
Air Blank	0.000	03:52	
Air Blank	0.000	03:54	
Subject Sample	# 2 0.123	03:57	Volume Not Met
Breath Volume	0.941	03:57	
Air Blank	0.000	03:58	
Control Test	0.080	03:58	
Air Blank	0.000	03:59	
Diagnostics Check	OK	03:59	

Data Download Date/Time: 01/31/2016 14:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DRISCOLL,KENNETH,B
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: WILDE,MATTHEW
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 12/22/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 19:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:04

SUBJECT'S LAST NAME: JACOBS

DATE/TIME OF TEST: 01/05/2016 20:31:26

Diagnostics Check	OK	20:24
Air Blank	0.000	20:25
Control Test	0.080	20:25
Air Blank	0.000	20:26
Subject Sample # 1	0.260	20:26
Breath Volume	2.683	20:26
Air Blank	0.000	20:27
Air Blank	0.000	20:29
Subject Sample # 2	0.261	20:29
Breath Volume	3.113	20:29
Air Blank	0.000	20:30
Control Test	0.079	20:30
Air Blank	0.000	20:31
Diagnostics Check	OK	20:31

Data Download Date/Time: 01/26/2016 10:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PETERSEN,CUTLER,W
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: PETERSEN,CULTER
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 12/22/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:53

SUBJECT'S LAST NAME: MONET

DATE/TIME OF TEST: 01/09/2016 05:28:55

Diagnostics Check	OK	05:22
Air Blank	0.000	05:23
Control Test	0.078	05:23
Air Blank	0.000	05:23
Subject Sample # 1	0.189	05:24
Breath Volume	2.679	05:24
Air Blank	0.000	05:25
Air Blank	0.000	05:26
Subject Sample # 2	0.183	05:27
Breath Volume	2.257	05:27
Air Blank	0.000	05:28
Control Test	0.078	05:28
Air Blank	0.000	05:28
Diagnostics Check	OK	05:28

Data Download Date/Time: 01/26/2016 10:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KEEL,JESSICA,M
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: KEEL,JESSICA
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 12/22/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:12

SUBJECT'S LAST NAME: BOUNTHAVILAY

DATE/TIME OF TEST: 01/15/2016 03:44:30

Diagnostics Check	OK	03:34
Air Blank	0.000	03:35
Control Test	0.079	03:35
Air Blank	0.000	03:36
Subject Sample # 1	0.124	03:38
Breath Volume	2.031	03:38
Air Blank	0.000	03:39
Air Blank	0.000	03:41
Subject Sample # 2	0.120	03:42
Breath Volume	1.664	03:42
Air Blank	0.000	03:43
Control Test	0.075	03:43
Air Blank	0.000	03:44
Diagnostics Check	OK	03:44

Data Download Date/Time: 01/26/2016 10:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPEAKMAN,ROBERT,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: SPEAKMAN,ROBERT
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 12/22/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 10:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:32

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 01/23/2016 01:00:39

Diagnostics Check	OK	00:54
Air Blank	0.000	00:54
Control Test	0.079	00:55
Air Blank	0.000	00:55
Subject Sample # 1	0.090	00:56
Breath Volume	4.019	00:56
Air Blank	0.000	00:56
Air Blank	0.000	00:58
Subject Sample # 2	0.093	00:59
Breath Volume	2.132	00:59
Air Blank	0.000	00:59
Control Test	0.077	01:00
Air Blank	0.000	01:00
Diagnostics Check	OK	01:00

Data Download Date/Time: 01/26/2016 10:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOGGS,JUSTIN,E
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: BOGGS,JUSTIN
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 12/22/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 08:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:21

SUBJECT'S LAST NAME: MERRIWEATHER

DATE/TIME OF TEST: 01/24/2016 09:47:59

Diagnostics Check	OK	09:41
Air Blank	0.000	09:42
Control Test	0.079	09:42
Air Blank	0.000	09:43
Subject Sample # 1	0.259	09:43
Breath Volume	4.566	09:43
Air Blank	0.000	09:44
Air Blank	0.000	09:45
Subject Sample # 2	0.253	09:46
Breath Volume	4.730	09:46
Air Blank	0.000	09:47
Control Test	0.077	09:47
Air Blank	0.000	09:47
Diagnostics Check	OK	09:47

Data Download Date/Time: 01/26/2016 10:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,GLENN,H
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: LINDSEY,PASCO
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 01/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 06:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:50

SUBJECT'S LAST NAME: BOLIN

DATE/TIME OF TEST: 01/30/2016 08:18:07

Diagnostics Check	OK	08:10
Air Blank	0.000	08:11
Control Test	0.076	08:11
Air Blank	0.000	08:11
Subject Sample # 1	0.193	08:13
Breath Volume	2.285	08:13
Air Blank	0.000	08:14
Air Blank	0.000	08:16
Subject Sample # 2	0.183	08:16
Breath Volume	3.042	08:16
Air Blank	0.000	08:17
Control Test	0.075	08:17
Air Blank	0.000	08:18
Diagnostics Check	OK	08:18

Data Download Date/Time: 02/29/2016 12:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBINSON,ROBERT,S
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: ROBINSON,ROBERT
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 01/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:21

SUBJECT'S LAST NAME: LOYED

DATE/TIME OF TEST: 01/31/2016 21:51:03

Diagnostics Check	OK	21:41	
Air Blank	0.000	21:41	
Control Test	0.079	21:42	
Air Blank	0.000	21:42	
Subject Sample	# 1	21:45	No Sample Provided
Breath Volume	0.000	21:45	
Air Blank	0.000	21:46	
Air Blank	0.000	21:48	
Subject Sample	# 2	21:49	Subject Test Refused
Breath Volume	0.000	21:49	
Air Blank	0.000	21:50	
Control Test	0.078	21:50	
Air Blank	0.000	21:50	
Diagnostics Check	OK	21:51	

Data Download Date/Time: 02/29/2016 12:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000742

INSTRUMENT REGISTERED TO: VALPARAISO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: TONEY,CRAIG,P

BREATH TEST OPERATOR AGENCY: Valparaiso PD

ARRESTING OFFICER: MALONEY,JAMES

ARRESTING OFFICER AGENCY: Valparaiso PD

LAST AGENCY INSPECTION: 12/27/2015

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1698962

DRY GAS EXPIRY DATE: 06/22/2017

VIOLATION CODE: DUI

ARREST TIME: 04:21

GENDER: Male

OBSERVATION PERIOD BEGAN: 05:03

SUBJECT'S LAST NAME: ELSENHEIMER

DATE/TIME OF TEST: 01/06/2016 05:31:54

Diagnostics Check	OK	05:25
Air Blank	0.000	05:26
Control Test	0.078	05:26
Air Blank	0.000	05:26
Subject Sample # 1	0.203	05:27
Breath Volume	4.128	05:27
Air Blank	0.000	05:28
Air Blank	0.000	05:29
Subject Sample # 2	0.207	05:30
Breath Volume	3.804	05:30
Air Blank	0.000	05:31
Control Test	0.077	05:31
Air Blank	0.000	05:31
Diagnostics Check	OK	05:31

Data Download Date/Time: 01/30/2016 07:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000742

INSTRUMENT REGISTERED TO: VALPARAISO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNGE, DOGBE, F
BREATH TEST OPERATOR AGENCY: Eglin AFB PD
ARRESTING OFFICER: TURNER, THOMAS
ARRESTING OFFICER AGENCY: Eglin AFB PD
LAST AGENCY INSPECTION: 12/27/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1698962
DRY GAS EXPIRY DATE: 06/22/2017
VIOLATION CODE: DUI
ARREST TIME: 05:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:04

SUBJECT'S LAST NAME: YOUNGE

DATE/TIME OF TEST: 01/10/2016 06:44:10

Diagnostics Check	OK	06:32	
Air Blank	0.000	06:33	
Control Test	0.079	06:33	
Air Blank	0.000	06:34	
Subject Sample	# 1 0.210	06:37	Volume Not Met
Breath Volume	0.320	06:37	
Air Blank	0.000	06:38	
Air Blank	0.000	06:39	
Subject Sample	# 2	06:42	No Sample Provided
Breath Volume	0.000	06:42	
Air Blank	0.000	06:43	
Control Test	0.078	06:43	
Air Blank	0.000	06:44	
Diagnostics Check	OK	06:44	

Data Download Date/Time: 01/30/2016 07:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000742

INSTRUMENT REGISTERED TO: VALPARAISO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HANK,JANINE,M

BREATH TEST OPERATOR AGENCY: Eglin AFB PD

ARRESTING OFFICER: PARRINO,TODD

ARRESTING OFFICER AGENCY: Eglin AFB PD

LAST AGENCY INSPECTION: 12/27/2015

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1698962

DRY GAS EXPIRY DATE: 06/22/2017

VIOLATION CODE: DUI

ARREST TIME: 02:02

GENDER: Female

OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: MORGAN

DATE/TIME OF TEST: 01/30/2016 02:35:33

Diagnostics Check	OK	02:28
Air Blank	0.000	02:28
Control Test	0.078	02:29
Air Blank	0.000	02:29
Subject Sample # 1	0.143	02:30
Breath Volume	1.476	02:30
Air Blank	0.000	02:31
Air Blank	0.000	02:33
Subject Sample # 2	0.133	02:34
Breath Volume	2.457	02:34
Air Blank	0.000	02:34
Control Test	0.078	02:34
Air Blank	0.000	02:35
Diagnostics Check	OK	02:35

Data Download Date/Time: 01/30/2016 07:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000752

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LETSON,REBECCA,G
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: MOOREHEAD,DAVID
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 12/29/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 83550
DRY GAS EXPIRY DATE: 09/26/2018
VIOLATION CODE: DUI
ARREST TIME: 00:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:16

SUBJECT'S LAST NAME: STANSBERRY

DATE/TIME OF TEST: 01/01/2016 00:45:26

Diagnostics Check	OK	00:39
Air Blank	0.000	00:39
Control Test	0.080	00:40
Air Blank	0.000	00:40
Subject Sample # 1	0.145	00:40
Breath Volume	1.851	00:40
Air Blank	0.000	00:41
Air Blank	0.000	00:43
Subject Sample # 2	0.147	00:43
Breath Volume	2.445	00:43
Air Blank	0.000	00:44
Control Test	0.080	00:44
Air Blank	0.000	00:45
Diagnostics Check	OK	00:45

Data Download Date/Time: 01/22/2016 00:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000752

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,GARY,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: MILLER,GARY
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 12/29/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 83550
DRY GAS EXPIRY DATE: 09/26/2018
VIOLATION CODE: DUI
ARREST TIME: 00:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:13

SUBJECT'S LAST NAME: STOLBRAND

DATE/TIME OF TEST: 01/01/2016 01:39:22

Diagnostics Check	OK	01:33
Air Blank	0.000	01:33
Control Test	0.080	01:34
Air Blank	0.000	01:34
Subject Sample # 1	0.158	01:35
Breath Volume	2.644	01:35
Air Blank	0.000	01:35
Air Blank	0.000	01:37
Subject Sample # 2	0.149	01:37
Breath Volume	1.621	01:37
Air Blank	0.000	01:38
Control Test	0.080	01:38
Air Blank	0.000	01:39
Diagnostics Check	OK	01:39

Data Download Date/Time: 01/22/2016 00:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000752

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,MANUEL,A
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: FRETTS,TIM
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 12/29/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 83550
DRY GAS EXPIRY DATE: 09/26/2018
VIOLATION CODE: DUI
ARREST TIME: 02:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:01

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 01/01/2016 03:45:47

Diagnostics Check	OK	03:39
Air Blank	0.000	03:39
Control Test	0.080	03:40
Air Blank	0.000	03:40
Subject Sample # 1	0.293	03:41
Breath Volume	2.863	03:41
Air Blank	0.000	03:41
Air Blank	0.000	03:43
Subject Sample # 2	0.285	03:43
Breath Volume	2.820	03:43
Air Blank	0.000	03:44
Control Test	0.079	03:45
Air Blank	0.000	03:45
Diagnostics Check	OK	03:45

Data Download Date/Time: 01/22/2016 00:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000752

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,MANUEL,A
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: HERNANDEZ,MANUEL
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 12/29/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 83550
DRY GAS EXPIRY DATE: 09/26/2018
VIOLATION CODE: DUI
ARREST TIME: 03:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: VENTI

DATE/TIME OF TEST: 01/01/2016 04:00:10

Diagnostics Check	OK	03:55	
Air Blank	0.000	03:56	
Control Test	0.079	03:56	
Air Blank	0.000	03:57	
Subject Sample	# 1	03:58	Subject Test Refused
Breath Volume	0.000	03:58	
Air Blank	0.000	03:59	
Control Test	0.079	03:59	
Air Blank	0.000	04:00	
Diagnostics Check	OK	04:00	

Data Download Date/Time: 01/22/2016 00:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000752

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,GARY,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: MILLER,GARY
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 12/29/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 83550
DRY GAS EXPIRY DATE: 09/26/2018
VIOLATION CODE: DUI
ARREST TIME: 03:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:57

SUBJECT'S LAST NAME: SUDMAN

DATE/TIME OF TEST: 01/01/2016 04:31:49

Diagnostics Check	OK	04:20	
Air Blank	0.000	04:20	
Control Test	0.080	04:20	
Air Blank	0.000	04:21	
Subject Sample	# 1 0.176	04:24	Volume Not Met
Breath Volume	0.437	04:24	
Air Blank	0.000	04:25	
Air Blank	0.000	04:27	
Subject Sample	# 2 0.172	04:30	Volume Not Met
Breath Volume	0.359	04:30	
Air Blank	0.000	04:30	
Control Test	0.078	04:31	
Air Blank	0.000	04:31	
Diagnostics Check	OK	04:31	

Data Download Date/Time: 01/22/2016 00:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000752

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LETSON,REBECCA,G
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: COLON,ARNALDO
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 12/29/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 83550
DRY GAS EXPIRY DATE: 09/26/2018
VIOLATION CODE: DUI
ARREST TIME: 03:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:04

SUBJECT'S LAST NAME: SANTINI ARRUBARRENA

DATE/TIME OF TEST: 01/01/2016 04:53:49

Diagnostics Check	OK	04:47
Air Blank	0.000	04:47
Control Test	0.079	04:47
Air Blank	0.000	04:48
Subject Sample # 1	0.125	04:49
Breath Volume	2.078	04:49
Air Blank	0.000	04:50
Air Blank	0.000	04:51
Subject Sample # 2	0.137	04:52
Breath Volume	2.089	04:52
Air Blank	0.000	04:52
Control Test	0.079	04:53
Air Blank	0.000	04:53
Diagnostics Check	OK	04:53

Data Download Date/Time: 01/22/2016 00:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000752

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DRIVER,JOSHUA,L
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: DRIVER,JOSHUA
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 12/29/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 83550
DRY GAS EXPIRY DATE: 09/26/2018
VIOLATION CODE: DUI
ARREST TIME: 02:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:54

SUBJECT'S LAST NAME: DOMINY

DATE/TIME OF TEST: 01/03/2016 03:21:27

Diagnostics Check	OK	03:15
Air Blank	0.000	03:15
Control Test	0.081	03:15
Air Blank	0.000	03:16
Subject Sample # 1	0.151	03:16
Breath Volume	1.382	03:16
Air Blank	0.000	03:17
Air Blank	0.000	03:19
Subject Sample # 2	0.169	03:19
Breath Volume	2.066	03:19
Air Blank	0.000	03:20
Control Test	0.080	03:20
Air Blank	0.000	03:21
Diagnostics Check	OK	03:21

Data Download Date/Time: 01/22/2016 00:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000752

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GAGO-
CHUQUILLANQUI,ALBERTO,E
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: GAGO-
CHUQUILLANQUI,ALBERTO
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 12/29/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 83550
DRY GAS EXPIRY DATE: 09/26/2018

VIOLATION CODE: DUI
ARREST TIME: 23:05
GENDER: Male

OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: EASTWOOD

DATE/TIME OF TEST: 01/04/2016 23:53:56

Diagnostics Check	OK	23:51	
Air Blank	0.000	23:51	
Control Test	0.081	23:51	
Air Blank	0.000	23:52	
Subject Sample	# 1	23:52	Subject Test Refused
Breath Volume	0.000	23:52	
Air Blank	0.000	23:53	
Control Test	0.081	23:53	
Air Blank	0.000	23:53	
Diagnostics Check	OK	23:53	

Data Download Date/Time: 01/22/2016 00:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000752

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LETSON,REBECCA,G
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: LETSON,REBECCA
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 12/29/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 83550
DRY GAS EXPIRY DATE: 09/26/2018
VIOLATION CODE: DUI
ARREST TIME: 00:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:21

SUBJECT'S LAST NAME: BLANK

DATE/TIME OF TEST: 01/06/2016 00:52:44

Diagnostics Check	OK	00:47	
Air Blank	0.000	00:47	
Control Test	0.081	00:48	
Air Blank	0.000	00:48	
Subject Sample	# 1	00:50	Subject Test Refused
Breath Volume	0.167	00:50	
Air Blank	0.000	00:51	
Control Test	0.081	00:52	
Air Blank	0.000	00:52	
Diagnostics Check	OK	00:52	

Data Download Date/Time: 01/22/2016 00:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000752

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRAWFORD,RICHARD,S
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: CRAWFORD,RICHARD
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 12/29/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 83550
DRY GAS EXPIRY DATE: 09/26/2018
VIOLATION CODE: DUI
ARREST TIME: 12:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:10

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 01/06/2016 13:38:19

Diagnostics Check	OK	13:32
Air Blank	0.000	13:32
Control Test	0.081	13:33
Air Blank	0.000	13:33
Subject Sample # 1	0.000	13:34
Breath Volume	1.894	13:34
Air Blank	0.000	13:34
Air Blank	0.000	13:36
Subject Sample # 2	0.000	13:36
Breath Volume	2.054	13:36
Air Blank	0.000	13:37
Control Test	0.081	13:37
Air Blank	0.000	13:38
Diagnostics Check	OK	13:38

Data Download Date/Time: 01/22/2016 00:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000752

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,GARY,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: MILLER,GARY
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 12/29/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 83550
DRY GAS EXPIRY DATE: 09/26/2018
VIOLATION CODE: DUI
ARREST TIME: 16:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:26

SUBJECT'S LAST NAME: TURNER

DATE/TIME OF TEST: 01/06/2016 17:52:46

Diagnostics Check	OK	17:46
Air Blank	0.000	17:47
Control Test	0.081	17:47
Air Blank	0.000	17:47
Subject Sample # 1	0.079	17:48
Breath Volume	2.527	17:48
Air Blank	0.000	17:48
Air Blank	0.000	17:50
Subject Sample # 2	0.082	17:51
Breath Volume	3.082	17:51
Air Blank	0.000	17:51
Control Test	0.081	17:52
Air Blank	0.000	17:52
Diagnostics Check	OK	17:52

Data Download Date/Time: 01/22/2016 00:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000752

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LETSON,REBECCA,G
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: LETSON,REBECCA
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 12/29/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 83550
DRY GAS EXPIRY DATE: 09/26/2018
VIOLATION CODE: DUI
ARREST TIME: 01:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:01

SUBJECT'S LAST NAME: EDWARDS

DATE/TIME OF TEST: 01/08/2016 02:39:01

Diagnostics Check	OK	02:32
Air Blank	0.000	02:32
Control Test	0.081	02:33
Air Blank	0.000	02:33
Subject Sample # 1	0.218	02:34
Breath Volume	1.933	02:34
Air Blank	0.000	02:35
Air Blank	0.000	02:36
Subject Sample # 2	0.224	02:37
Breath Volume	3.273	02:37
Air Blank	0.000	02:38
Control Test	0.078	02:38
Air Blank	0.000	02:38
Diagnostics Check	OK	02:39

Data Download Date/Time: 01/22/2016 00:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000752

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GAGO-
CHUQUILLANQUI,ALBERTO,E
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: GAGO-
CHUQUILLANQUI,ALBERTO
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 12/29/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 83550
DRY GAS EXPIRY DATE: 09/26/2018

VIOLATION CODE: DUI
ARREST TIME: 23:16
GENDER: Male

OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: MARCUM

DATE/TIME OF TEST: 01/09/2016 00:43:39

Diagnostics Check	OK	00:36
Air Blank	0.000	00:37
Control Test	0.081	00:37
Air Blank	0.000	00:37
Subject Sample # 1	0.102	00:38
Breath Volume	1.933	00:38
Air Blank	0.000	00:39
Air Blank	0.000	00:41
Subject Sample # 2	0.092	00:42
Breath Volume	1.621	00:42
Air Blank	0.000	00:42
Control Test	0.079	00:43
Air Blank	0.000	00:43
Diagnostics Check	OK	00:43

Data Download Date/Time: 01/22/2016 00:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000752

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GAGO-
CHUQUILLANQUI,ALBERTO,E
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: GAGO-
CHUQUILLANQUI,ALBERTO
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 12/29/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 83550
DRY GAS EXPIRY DATE: 09/26/2018

VIOLATION CODE: DUI
ARREST TIME: 23:23
GENDER: Male

OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: IRWIN

DATE/TIME OF TEST: 01/10/2016 00:28:49

Diagnostics Check	OK	00:22
Air Blank	0.000	00:22
Control Test	0.080	00:23
Air Blank	0.000	00:23
Subject Sample # 1	0.235	00:24
Breath Volume	1.785	00:24
Air Blank	0.000	00:24
Air Blank	0.000	00:26
Subject Sample # 2	0.224	00:27
Breath Volume	1.644	00:27
Air Blank	0.000	00:27
Control Test	0.079	00:28
Air Blank	0.000	00:28
Diagnostics Check	OK	00:28

Data Download Date/Time: 01/22/2016 00:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000752

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GERNERT,JOHN,D
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: GERNERT,JOHN
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 12/29/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 83550
DRY GAS EXPIRY DATE: 09/26/2018
VIOLATION CODE: DUI
ARREST TIME: 22:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: BURKHART

DATE/TIME OF TEST: 01/13/2016 00:29:04

Diagnostics Check	OK	00:16	
Air Blank	0.000	00:17	
Control Test	0.081	00:17	
Air Blank	0.000	00:17	
Subject Sample	# 1 0.126	00:21	Volume Not Met
Breath Volume	0.195	00:21	
Air Blank	0.000	00:22	
Air Blank	0.000	00:23	
Subject Sample	# 2 0.157	00:24	No .020 agreement
Breath Volume	1.726	00:24	
Air Blank	0.000	00:25	
Air Blank	0.000	00:26	
Subject Sample	# 3 0.132	00:27	No .020 agreement
Breath Volume	2.113	00:27	
Air Blank	0.000	00:28	
Control Test	0.075	00:28	
Air Blank	0.000	00:28	
Diagnostics Check	OK	00:29	

Data Download Date/Time: 01/22/2016 00:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000752

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LETSON,REBECCA,G
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: MOOREHEAD,DAVID
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 12/29/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 83550
DRY GAS EXPIRY DATE: 09/26/2018
VIOLATION CODE: DUI
ARREST TIME: 00:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:57

SUBJECT'S LAST NAME: ABRAHAMSON COLON

DATE/TIME OF TEST: 01/13/2016 01:38:05

Diagnostics Check	OK	01:26	
Air Blank	0.000	01:27	
Control Test	0.080	01:27	
Air Blank	0.000	01:28	
Subject Sample	# 1 0.168	01:31	Volume Not Met
Breath Volume	0.066	01:31	
Air Blank	0.000	01:32	
Air Blank	0.000	01:33	
Subject Sample	# 2 0.163	01:36	Volume Not Met
Breath Volume	0.070	01:36	
Air Blank		01:37	Purge Fail
Air Blank	0.000	01:38	

Data Download Date/Time: 01/22/2016 00:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000752

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LETSON,REBECCA,G
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: MOOREHEAD,DAVID
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 12/29/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 83550
DRY GAS EXPIRY DATE: 09/26/2018
VIOLATION CODE: DUI
ARREST TIME: 00:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:57

SUBJECT'S LAST NAME: ABRAHAMSON COLON

DATE/TIME OF TEST: 01/13/2016 01:44:37

Diagnostics Check	OK	01:40	
Air Blank	0.000	01:40	
Control Test	0.077	01:41	
Air Blank	0.000	01:41	
Subject Sample	# 1	01:42	Subject Test Refused
Breath Volume	0.328	01:42	
Air Blank	0.000	01:43	
Control Test	0.078	01:44	
Air Blank	0.000	01:44	
Diagnostics Check	OK	01:44	

Data Download Date/Time: 01/22/2016 00:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000752

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCFARLAND,MICHAEL,D
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: MCFARLAND,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 12/29/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1766638
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 16:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:25

SUBJECT'S LAST NAME: CHAMBERS

DATE/TIME OF TEST: 01/13/2016 18:01:22

Diagnostics Check	OK	17:53
Air Blank	0.000	17:54
Control Test	0.081	17:54
Air Blank	0.000	17:55
Subject Sample # 1	0.174	17:56
Breath Volume	1.789	17:56
Air Blank	0.000	17:56
Air Blank	0.000	17:58
Subject Sample # 2	0.165	17:59
Breath Volume	2.949	17:59
Air Blank	0.000	18:00
Control Test	0.079	18:00
Air Blank	0.000	18:01
Diagnostics Check	OK	18:01

Data Download Date/Time: 01/22/2016 00:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000752

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,GARY,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: MILLER,GARY
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 12/29/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 83550
DRY GAS EXPIRY DATE: 09/26/2018
VIOLATION CODE: DUI
ARREST TIME: 00:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:27

SUBJECT'S LAST NAME: LEWIS

DATE/TIME OF TEST: 01/14/2016 01:53:26

Diagnostics Check	OK	01:47
Air Blank	0.000	01:47
Control Test	0.081	01:47
Air Blank	0.000	01:48
Subject Sample # 1	0.070	01:49
Breath Volume	1.796	01:49
Air Blank	0.000	01:49
Air Blank	0.000	01:51
Subject Sample # 2	0.071	01:51
Breath Volume	1.464	01:51
Air Blank	0.000	01:52
Control Test	0.081	01:52
Air Blank	0.000	01:53
Diagnostics Check	OK	01:53

Data Download Date/Time: 01/22/2016 00:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000752

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SERRANO,ELI,
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: CRAIG,J
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 12/29/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 83550
DRY GAS EXPIRY DATE: 09/26/2018
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: WOSYLUK

DATE/TIME OF TEST: 01/14/2016 02:38:23

Diagnostics Check	OK	02:31
Air Blank	0.000	02:31
Control Test	0.081	02:31
Air Blank	0.000	02:32
Subject Sample # 1	0.136	02:33
Breath Volume	2.320	02:33
Air Blank	0.000	02:34
Air Blank	0.000	02:36
Subject Sample # 2	0.138	02:36
Breath Volume	2.003	02:36
Air Blank	0.000	02:37
Control Test	0.079	02:37
Air Blank	0.000	02:38
Diagnostics Check	OK	02:38

Data Download Date/Time: 01/22/2016 00:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000752

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURLESON,BILLY,M
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: BURLESON,BILLY
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 12/29/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 83550
DRY GAS EXPIRY DATE: 09/26/2018
VIOLATION CODE: DUI
ARREST TIME: 23:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:48

SUBJECT'S LAST NAME: WAGNER

DATE/TIME OF TEST: 01/15/2016 00:18:54

Diagnostics Check	OK	00:08
Air Blank	0.000	00:09
Control Test	0.081	00:09
Air Blank	0.000	00:09
Subject Sample # 1	0.118	00:12
Breath Volume	2.261	00:12
Air Blank	0.000	00:13
Air Blank	0.000	00:15
Subject Sample # 2	0.125	00:17
Breath Volume	3.187	00:17
Air Blank	0.000	00:17
Control Test	0.081	00:18
Air Blank	0.000	00:18
Diagnostics Check	OK	00:18

Data Download Date/Time: 01/22/2016 00:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000752

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SERRANO,ELI,
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: SERRANO,E
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 12/29/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 83550
DRY GAS EXPIRY DATE: 09/26/2018
VIOLATION CODE: DUI
ARREST TIME: 02:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: DEVILLE

DATE/TIME OF TEST: 01/15/2016 03:17:58

Diagnostics Check	OK	03:11
Air Blank	0.000	03:11
Control Test	0.081	03:12
Air Blank	0.000	03:12
Subject Sample # 1	0.158	03:13
Breath Volume	3.996	03:13
Air Blank	0.000	03:13
Air Blank	0.000	03:15
Subject Sample # 2	0.152	03:16
Breath Volume	3.542	03:16
Air Blank	0.000	03:17
Control Test	0.079	03:17
Air Blank	0.000	03:17
Diagnostics Check	OK	03:17

Data Download Date/Time: 01/22/2016 00:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000752

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COUGHLIN,MICHAEL,P
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: MOOREHEAD,DAVID
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 12/29/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 83550
DRY GAS EXPIRY DATE: 09/26/2018
VIOLATION CODE: DUI
ARREST TIME: 00:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:52

SUBJECT'S LAST NAME: LIPA

DATE/TIME OF TEST: 01/16/2016 01:25:07

Diagnostics Check	OK	01:16	
Air Blank	0.000	01:17	
Control Test	0.081	01:17	
Air Blank	0.000	01:17	
Subject Sample	# 1	01:21	No Sample Provided
Breath Volume	0.000	01:21	
Air Blank	0.000	01:21	
Air Blank	0.000	01:23	
Subject Sample	# 2	01:23	Subject Test Refused
Breath Volume	0.000	01:23	
Air Blank	0.000	01:24	
Control Test	0.081	01:24	
Air Blank	0.000	01:25	
Diagnostics Check	OK	01:25	

Data Download Date/Time: 01/22/2016 00:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000752

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOBOSIEWICZ, ANTHONY, N
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: DOBOSIEWICZ, ANTHONY
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 12/29/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 83550
DRY GAS EXPIRY DATE: 09/26/2018
VIOLATION CODE: DUI
ARREST TIME: 21:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: DAVIDSON

DATE/TIME OF TEST: 01/17/2016 22:15:06

Diagnostics Check	OK	22:08
Air Blank	0.000	22:08
Control Test	0.081	22:08
Air Blank	0.000	22:09
Subject Sample # 1	0.192	22:10
Breath Volume	1.847	22:10
Air Blank	0.000	22:11
Air Blank	0.000	22:12
Subject Sample # 2	0.188	22:13
Breath Volume	2.496	22:13
Air Blank	0.000	22:14
Control Test	0.080	22:14
Air Blank	0.000	22:15
Diagnostics Check	OK	22:15

Data Download Date/Time: 01/22/2016 00:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000752

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERNANDEZ,ALAIN,
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: FERNANDEZ,ALAIN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 12/29/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 83550
DRY GAS EXPIRY DATE: 09/26/2018
VIOLATION CODE: DUI
ARREST TIME: 23:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:43

SUBJECT'S LAST NAME: BOLDEN

DATE/TIME OF TEST: 01/19/2016 01:10:34

Diagnostics Check	OK	01:04
Air Blank	0.000	01:04
Control Test	0.082	01:04
Air Blank	0.000	01:05
Subject Sample # 1	0.147	01:06
Breath Volume	1.734	01:06
Air Blank	0.000	01:06
Air Blank	0.000	01:08
Subject Sample # 2	0.145	01:08
Breath Volume	1.562	01:08
Air Blank	0.000	01:09
Control Test	0.081	01:09
Air Blank	0.000	01:10
Diagnostics Check	OK	01:10

Data Download Date/Time: 01/22/2016 00:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000752

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: UMANA,RAUL,J
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: UMANA,RAUL
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 12/29/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 83550
DRY GAS EXPIRY DATE: 09/26/2018
VIOLATION CODE: DUI
ARREST TIME: 23:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:32

SUBJECT'S LAST NAME: DICKS

DATE/TIME OF TEST: 01/19/2016 23:59:06

Diagnostics Check	OK	23:52
Air Blank	0.000	23:53
Control Test	0.081	23:53
Air Blank	0.000	23:54
Subject Sample # 1	0.134	23:54
Breath Volume	1.152	23:54
Air Blank	0.000	23:55
Air Blank	0.000	23:57
Subject Sample # 2	0.133	23:57
Breath Volume	1.128	23:57
Air Blank	0.000	23:58
Control Test	0.081	23:58
Air Blank	0.000	23:59
Diagnostics Check	OK	23:59

Data Download Date/Time: 01/22/2016 00:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000752

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRAWFORD,RICHARD,S
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: CRAWFORD,RICHARD
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 12/29/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 83550
DRY GAS EXPIRY DATE: 09/26/2018
VIOLATION CODE: DUI
ARREST TIME: 16:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:15

SUBJECT'S LAST NAME: COLLINS

DATE/TIME OF TEST: 01/20/2016 17:48:24

Diagnostics Check	OK	17:41
Air Blank	0.000	17:42
Control Test	0.081	17:42
Air Blank	0.000	17:43
Subject Sample # 1	0.278	17:43
Breath Volume	1.746	17:43
Air Blank	0.000	17:44
Air Blank	0.000	17:45
Subject Sample # 2	0.268	17:46
Breath Volume	1.828	17:46
Air Blank	0.000	17:47
Control Test	0.081	17:47
Air Blank	0.000	17:48
Diagnostics Check	OK	17:48

Data Download Date/Time: 01/22/2016 00:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000752

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,GARY,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: MILLER,GARY
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 12/29/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 83550
DRY GAS EXPIRY DATE: 09/26/2018
VIOLATION CODE: DUI
ARREST TIME: 22:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:17

SUBJECT'S LAST NAME: SUOZZI

DATE/TIME OF TEST: 01/21/2016 23:45:43

Diagnostics Check	OK	23:39
Air Blank	0.000	23:39
Control Test	0.082	23:40
Air Blank	0.000	23:40
Subject Sample # 1	0.000	23:41
Breath Volume	1.777	23:41
Air Blank	0.000	23:41
Air Blank	0.000	23:43
Subject Sample # 2	0.000	23:44
Breath Volume	1.992	23:44
Air Blank	0.000	23:44
Control Test	0.081	23:45
Air Blank	0.000	23:45
Diagnostics Check	OK	23:45

Data Download Date/Time: 01/22/2016 00:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUNT,CHARLES,R
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: HUNT,CHARLES
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 12/24/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 05:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:45

SUBJECT'S LAST NAME: RODRIGUEZ PACHECO

DATE/TIME OF TEST: 01/02/2016 06:20:22

Diagnostics Check	OK	06:13
Air Blank	0.000	06:13
Control Test	0.079	06:14
Air Blank	0.000	06:14
Subject Sample # 1	0.129	06:15
Breath Volume	1.765	06:15
Air Blank	0.000	06:16
Air Blank	0.000	06:18
Subject Sample # 2	0.119	06:18
Breath Volume	1.406	06:18
Air Blank	0.000	06:19
Control Test	0.077	06:19
Air Blank	0.000	06:20
Diagnostics Check	OK	06:20

Data Download Date/Time: 01/22/2016 10:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOFMANN,TRAVIS,A
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: HOFMANN,TRAVIS
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 12/24/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: WESTOVER

DATE/TIME OF TEST: 01/04/2016 04:00:32

Diagnostics Check	OK	03:53
Air Blank	0.000	03:54
Control Test	0.080	03:54
Air Blank	0.000	03:55
Subject Sample # 1	0.153	03:56
Breath Volume	1.789	03:56
Air Blank	0.000	03:56
Air Blank	0.000	03:58
Subject Sample # 2	0.152	03:58
Breath Volume	1.695	03:58
Air Blank	0.000	03:59
Control Test	0.080	03:59
Air Blank	0.000	04:00
Diagnostics Check	OK	04:00

Data Download Date/Time: 01/22/2016 10:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRICKER,JEFFREY,S
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: YANCY,JAMES
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 12/24/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 15:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:11

SUBJECT'S LAST NAME: SEIDL

DATE/TIME OF TEST: 01/05/2016 17:47:29

Diagnostics Check	OK	17:39	
Air Blank	0.000	17:39	
Control Test	0.079	17:39	
Air Blank	0.000	17:40	
Subject Sample	# 1	17:43	No Sample Provided
Breath Volume	0.000	17:43	
Air Blank	0.000	17:43	
Air Blank	0.000	17:45	
Subject Sample	# 2	17:46	Subject Test Refused
Breath Volume	0.000	17:46	
Air Blank	0.000	17:46	
Control Test	0.080	17:46	
Air Blank	0.000	17:47	
Diagnostics Check	OK	17:47	

Data Download Date/Time: 01/22/2016 10:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPBELL,ROBERT,E
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: DASHER,CHRISTOPHER
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 12/24/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:57

SUBJECT'S LAST NAME: ROBERTS

DATE/TIME OF TEST: 01/06/2016 04:24:16

Diagnostics Check	OK	04:17
Air Blank	0.000	04:17
Control Test	0.079	04:17
Air Blank	0.000	04:18
Subject Sample # 1	0.319	04:19
Breath Volume	2.503	04:19
Air Blank	0.000	04:20
Air Blank	0.000	04:21
Subject Sample # 2	0.318	04:22
Breath Volume	1.773	04:22
Air Blank	0.000	04:23
Control Test	0.077	04:23
Air Blank	0.000	04:24
Diagnostics Check	OK	04:24

Data Download Date/Time: 01/22/2016 10:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPBELL,ROBERT,E
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: DASHER,CHRISTOPHER
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 12/24/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:07

SUBJECT'S LAST NAME: BARR

DATE/TIME OF TEST: 01/08/2016 22:35:32

Diagnostics Check	OK	22:28
Air Blank	0.000	22:28
Control Test	0.079	22:28
Air Blank	0.000	22:29
Subject Sample # 1	0.155	22:30
Breath Volume	1.847	22:30
Air Blank	0.000	22:31
Air Blank	0.000	22:33
Subject Sample # 2	0.147	22:33
Breath Volume	1.550	22:33
Air Blank	0.000	22:34
Control Test	0.076	22:34
Air Blank	0.000	22:35
Diagnostics Check	OK	22:35

Data Download Date/Time: 01/22/2016 10:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRIVETTE,CHRIS,A
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: PRIVETTE,CHRIS
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 12/24/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 15:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:09

SUBJECT'S LAST NAME: JORDAN

DATE/TIME OF TEST: 01/10/2016 17:43:38

Diagnostics Check	OK	17:36
Air Blank	0.000	17:37
Control Test	0.079	17:37
Air Blank	0.000	17:38
Subject Sample # 1	0.000	17:39
Breath Volume	1.792	17:39
Air Blank	0.000	17:39
Air Blank	0.000	17:41
Subject Sample # 2	0.000	17:42
Breath Volume	1.625	17:42
Air Blank	0.000	17:42
Control Test	0.078	17:43
Air Blank	0.000	17:43
Diagnostics Check	OK	17:43

Data Download Date/Time: 01/22/2016 10:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUNT,CHARLES,R
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: HUNT,CHARLES
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 12/24/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: ENGEL

DATE/TIME OF TEST: 01/13/2016 03:42:43

Diagnostics Check	OK	03:36
Air Blank	0.000	03:36
Control Test	0.079	03:37
Air Blank	0.000	03:37
Subject Sample # 1	0.128	03:38
Breath Volume	2.050	03:38
Air Blank	0.000	03:38
Air Blank	0.000	03:40
Subject Sample # 2	0.131	03:41
Breath Volume	1.800	03:41
Air Blank	0.000	03:41
Control Test	0.080	03:42
Air Blank	0.000	03:42
Diagnostics Check	OK	03:42

Data Download Date/Time: 01/22/2016 10:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICHARDSON,MICHAEL,R
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: RICHARDSON,MICHAEL
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 12/24/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:23

SUBJECT'S LAST NAME: JEREMIAH

DATE/TIME OF TEST: 01/14/2016 23:55:13

Diagnostics Check	OK	23:48
Air Blank	0.000	23:49
Control Test	0.079	23:49
Air Blank	0.000	23:50
Subject Sample # 1	0.133	23:50
Breath Volume	2.203	23:50
Air Blank	0.000	23:51
Air Blank	0.000	23:53
Subject Sample # 2	0.125	23:53
Breath Volume	1.460	23:53
Air Blank	0.000	23:54
Control Test	0.079	23:54
Air Blank	0.000	23:55
Diagnostics Check	OK	23:55

Data Download Date/Time: 01/22/2016 10:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOSTICK,JESSE,W
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: WENDT,HENDRIK
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 12/24/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: REYNOLDS

DATE/TIME OF TEST: 01/15/2016 03:16:58

Diagnostics Check	OK	03:10
Air Blank	0.000	03:11
Control Test	0.079	03:11
Air Blank	0.000	03:11
Subject Sample # 1	0.211	03:12
Breath Volume	2.757	03:12
Air Blank	0.000	03:13
Air Blank	0.000	03:14
Subject Sample # 2	0.192	03:15
Breath Volume	2.015	03:15
Air Blank	0.000	03:16
Control Test	0.078	03:16
Air Blank	0.000	03:16
Diagnostics Check	OK	03:16

Data Download Date/Time: 01/22/2016 10:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNS, WILLIAM, A
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: MOORE, GARY
ARRESTING OFFICER AGENCY: High Springs PD
LAST AGENCY INSPECTION: 12/24/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 01/16/2016 03:21:33

Diagnostics Check	OK	03:10	
Air Blank	0.000	03:10	
Control Test	0.079	03:10	
Air Blank	0.000	03:11	
Subject Sample	# 1 0.131	03:14	Volume Not Met
Breath Volume	0.082	03:14	
Air Blank	0.000	03:15	
Air Blank	0.000	03:16	
Subject Sample	# 2 0.000	03:19	Volume Not Met
Breath Volume	0.015	03:19	
Air Blank	0.000	03:20	
Control Test	0.078	03:20	
Air Blank	0.000	03:21	
Diagnostics Check	OK	03:21	

Data Download Date/Time: 01/22/2016 10:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOSTICK,MEGAN,K
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: WENDT,HENDRICK
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 12/24/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: HARVEY

DATE/TIME OF TEST: 01/17/2016 01:28:04

Diagnostics Check	OK	01:25	
Air Blank	0.000	01:25	
Control Test	0.079	01:25	
Air Blank	0.000	01:26	
Subject Sample	# 1	01:26	Subject Test Refused
Breath Volume	0.000	01:26	
Air Blank	0.000	01:27	
Control Test	0.078	01:27	
Air Blank	0.000	01:27	
Diagnostics Check	OK	01:28	

Data Download Date/Time: 01/22/2016 10:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOFMANN, TRAVIS, A
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: HOFMANN, TRAVIS
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 12/24/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: SHADRIX

DATE/TIME OF TEST: 01/17/2016 22:24:09

Diagnostics Check	OK	22:17
Air Blank	0.000	22:18
Control Test	0.079	22:18
Air Blank	0.000	22:19
Subject Sample # 1	0.180	22:19
Breath Volume	3.781	22:19
Air Blank	0.000	22:20
Air Blank	0.000	22:22
Subject Sample # 2	0.178	22:22
Breath Volume	3.359	22:22
Air Blank	0.000	22:23
Control Test	0.079	22:23
Air Blank	0.000	22:24
Diagnostics Check	OK	22:24

Data Download Date/Time: 01/22/2016 10:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRIVETTE,CHRIS,A
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: PRIVETTE,CHRIS
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 01/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 13:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:21

SUBJECT'S LAST NAME: BISACKY

DATE/TIME OF TEST: 01/22/2016 14:48:47

Diagnostics Check	OK	14:41	
Air Blank	0.000	14:42	
Control Test	0.079	14:42	
Air Blank	0.000	14:43	
Subject Sample # 1	0.000	14:44	
Breath Volume	3.632	14:44	
Air Blank	0.000	14:44	
Air Blank	0.000	14:46	
Subject Sample # 2	0.000	14:47	
Breath Volume	3.570	14:47	
Air Blank	0.000	14:47	
Control Test	0.078	14:47	
Air Blank		14:48	RFI Detect
Air Blank	0.000	14:48	

Data Download Date/Time: 02/26/2016 08:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCFARLAND,MICHAEL,D
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: MCFARLAND,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 01/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 16:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:45

SUBJECT'S LAST NAME: BARTHEL

DATE/TIME OF TEST: 01/23/2016 17:34:15

Diagnostics Check	OK	17:28
Air Blank	0.000	17:28
Control Test	0.080	17:28
Air Blank	0.000	17:29
Subject Sample # 1	0.000	17:29
Breath Volume	2.187	17:29
Air Blank	0.000	17:30
Air Blank	0.000	17:32
Subject Sample # 2	0.000	17:32
Breath Volume	2.628	17:32
Air Blank	0.000	17:33
Control Test	0.079	17:33
Air Blank	0.000	17:34
Diagnostics Check	OK	17:34

Data Download Date/Time: 02/26/2016 08:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOSTICK,JESSE,W
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: LAKE,JAMES
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 01/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:06

SUBJECT'S LAST NAME: RICHARDSON

DATE/TIME OF TEST: 01/24/2016 03:30:50

Diagnostics Check	OK	03:30	
Air Blank		03:30	RFI Detect
Air Blank	0.000	03:30	

Data Download Date/Time: 02/26/2016 08:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOSTICK,JESSE,W
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: LAKE,JAMES
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 01/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:06

SUBJECT'S LAST NAME: RICHARDSON

DATE/TIME OF TEST: 01/24/2016 03:41:24

Diagnostics Check	OK	03:31
Air Blank	0.000	03:32
Control Test	0.080	03:32
Air Blank	0.000	03:33
Subject Sample # 1	0.262	03:33
Breath Volume	3.507	03:33
Air Blank	0.000	03:34
Air Blank	0.000	03:36
Subject Sample # 2	0.234	03:36
Breath Volume	1.160	03:36
Air Blank	0.000	03:37
Air Blank	0.000	03:39
Subject Sample # 3	0.256	03:39
Breath Volume	3.093	03:39
Air Blank	0.000	03:40
Control Test	0.079	03:40
Air Blank	0.000	03:41
Diagnostics Check	OK	03:41

Data Download Date/Time: 02/26/2016 08:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUNT,CHARLES,R
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: HUNT,CHARLES
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 01/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:53

SUBJECT'S LAST NAME: SCHULBACH

DATE/TIME OF TEST: 01/27/2016 01:33:12

Diagnostics Check	OK	01:20	
Air Blank	0.000	01:20	
Control Test	0.079	01:20	
Air Blank	0.000	01:21	
Subject Sample	# 1 0.012	01:24	Volume Not Met
Breath Volume	0.066	01:24	
Air Blank	0.000	01:25	
Air Blank	0.000	01:26	
Subject Sample	# 2 0.009	01:27	
Breath Volume	1.476	01:27	
Air Blank	0.000	01:28	
Air Blank	0.000	01:30	
Subject Sample	# 3 0.008	01:31	
Breath Volume	1.562	01:31	
Air Blank	0.000	01:32	
Control Test	0.078	01:32	
Air Blank	0.000	01:33	
Diagnostics Check	OK	01:33	

Data Download Date/Time: 02/26/2016 08:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUNT,CHARLES,R
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: HUNT,CHARLES
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 01/22/2016

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:44

SUBJECT'S LAST NAME: AMER

DATE/TIME OF TEST: 01/30/2016 05:13:28

Diagnostics Check	OK	05:07
Air Blank	0.000	05:07
Control Test	0.079	05:07
Air Blank	0.000	05:08
Subject Sample # 1	0.188	05:08
Breath Volume	1.710	05:08
Air Blank	0.000	05:09
Air Blank	0.000	05:11
Subject Sample # 2	0.183	05:11
Breath Volume	1.878	05:11
Air Blank	0.000	05:12
Control Test	0.080	05:12
Air Blank	0.000	05:13
Diagnostics Check	OK	05:13

Data Download Date/Time: 02/26/2016 08:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000770

INSTRUMENT REGISTERED TO: HAMILTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARVEY,DAVID,N
BREATH TEST OPERATOR AGENCY: Jasper PD
ARRESTING OFFICER: HARVEY,DAVID
ARRESTING OFFICER AGENCY: Jasper PD
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06415080A2
DRY GAS EXPIRY DATE: 04/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:37

SUBJECT'S LAST NAME: OWENS

DATE/TIME OF TEST: 01/01/2016 03:05:01

Diagnostics Check	OK	02:58
Air Blank	0.000	02:58
Control Test	0.079	02:58
Air Blank	0.000	02:59
Subject Sample # 1	0.117	03:00
Breath Volume	3.148	03:00
Air Blank	0.000	03:01
Air Blank	0.000	03:02
Subject Sample # 2	0.108	03:03
Breath Volume	2.253	03:03
Air Blank	0.000	03:04
Control Test	0.078	03:04
Air Blank	0.000	03:04
Diagnostics Check	OK	03:05

Data Download Date/Time: 01/01/2016 21:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000770

INSTRUMENT REGISTERED TO: HAMILTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWNFIELD, CECIL, L
BREATH TEST OPERATOR AGENCY: White Springs PD
ARRESTING OFFICER: BROWNFIELD, CECIL
ARRESTING OFFICER AGENCY: White Springs PD
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06415080A2
DRY GAS EXPIRY DATE: 04/05/2017
VIOLATION CODE: DUI
ARREST TIME: 11:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:50

SUBJECT'S LAST NAME: CARBONE

DATE/TIME OF TEST: 01/04/2016 13:18:56

Diagnostics Check	OK	13:12
Air Blank	0.000	13:13
Control Test	0.080	13:13
Air Blank	0.000	13:13
Subject Sample # 1	0.166	13:14
Breath Volume	1.945	13:14
Air Blank	0.000	13:15
Air Blank	0.000	13:16
Subject Sample # 2	0.162	13:17
Breath Volume	2.222	13:17
Air Blank	0.000	13:18
Control Test	0.078	13:18
Air Blank	0.000	13:18
Diagnostics Check	OK	13:18

Data Download Date/Time: 01/31/2016 21:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000770

INSTRUMENT REGISTERED TO: HAMILTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARVEY,DAVID,A
BREATH TEST OPERATOR AGENCY: Jasper PD
ARRESTING OFFICER: HARVEY,DAVID
ARRESTING OFFICER AGENCY: Jasper PD
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06415080A2
DRY GAS EXPIRY DATE: 04/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:36

SUBJECT'S LAST NAME: CLARIDY

DATE/TIME OF TEST: 01/10/2016 00:00:46

Diagnostics Check	OK	23:59	
Air Blank		00:00	RFI Detect
Air Blank	0.000	00:00	

Data Download Date/Time: 01/31/2016 21:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000770

INSTRUMENT REGISTERED TO: HAMILTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARVEY,DAVID,A
BREATH TEST OPERATOR AGENCY: Jasper PD
ARRESTING OFFICER: HARVEY,DAVID
ARRESTING OFFICER AGENCY: Jasper PD
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06415080A2
DRY GAS EXPIRY DATE: 04/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:36

SUBJECT'S LAST NAME: CLARIDY

DATE/TIME OF TEST: 01/10/2016 00:13:33

Diagnostics Check	OK	00:04
Air Blank	0.000	00:05
Control Test	0.079	00:05
Air Blank	0.000	00:05
Subject Sample # 1	0.104	00:08
Breath Volume	2.875	00:08
Air Blank	0.000	00:09
Air Blank	0.000	00:10
Subject Sample # 2	0.099	00:11
Breath Volume	2.820	00:11
Air Blank	0.000	00:12
Control Test	0.077	00:12
Air Blank	0.000	00:13
Diagnostics Check	OK	00:13

Data Download Date/Time: 01/31/2016 21:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000770

INSTRUMENT REGISTERED TO: HAMILTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARVEY,DANTE,J
BREATH TEST OPERATOR AGENCY: Jasper PD
ARRESTING OFFICER: HARVEY,DAVID
ARRESTING OFFICER AGENCY: Jasper PD
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06415080A2
DRY GAS EXPIRY DATE: 04/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:31

SUBJECT'S LAST NAME: STAGGERS

DATE/TIME OF TEST: 01/31/2016 04:59:30

Diagnostics Check	OK	04:52
Air Blank	0.000	04:53
Control Test	0.080	04:53
Air Blank	0.000	04:54
Subject Sample # 1	0.175	04:54
Breath Volume	2.351	04:54
Air Blank	0.000	04:55
Air Blank	0.000	04:57
Subject Sample # 2	0.174	04:57
Breath Volume	2.175	04:57
Air Blank	0.000	04:58
Control Test	0.079	04:58
Air Blank	0.000	04:59
Diagnostics Check	OK	04:59

Data Download Date/Time: 01/31/2016 21:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COUNTS,RYAN,L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: COUNTS,RYAN
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG431502
DRY GAS EXPIRY DATE: 11/11/2016
VIOLATION CODE: DUI
ARREST TIME: 18:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:37

SUBJECT'S LAST NAME: KENEALY

DATE/TIME OF TEST: 01/01/2016 19:03:53

Diagnostics Check	OK	18:57
Air Blank	0.000	18:57
Control Test	0.080	18:58
Air Blank	0.000	18:58
Subject Sample # 1	0.086	18:59
Breath Volume	2.597	18:59
Air Blank	0.000	19:00
Air Blank	0.000	19:01
Subject Sample # 2	0.078	19:02
Breath Volume	1.261	19:02
Air Blank	0.000	19:02
Control Test	0.078	19:03
Air Blank	0.000	19:03
Diagnostics Check	OK	19:03

Data Download Date/Time: 01/08/2016 12:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VITALE,MICHAEL,L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: VITALE,MICHAEL
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG431502
DRY GAS EXPIRY DATE: 11/11/2016
VIOLATION CODE: DUI
ARREST TIME: 20:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:42

SUBJECT'S LAST NAME: DEJAYNES

DATE/TIME OF TEST: 01/03/2016 21:09:46

Diagnostics Check	OK	21:02
Air Blank	0.000	21:03
Control Test	0.080	21:03
Air Blank	0.000	21:04
Subject Sample # 1	0.000	21:04
Breath Volume	1.273	21:04
Air Blank	0.000	21:05
Air Blank	0.000	21:07
Subject Sample # 2	0.000	21:08
Breath Volume	1.894	21:08
Air Blank	0.000	21:08
Control Test	0.081	21:09
Air Blank	0.000	21:09
Diagnostics Check	OK	21:09

Data Download Date/Time: 01/08/2016 12:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAHL,RANDY,J
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: HARTBARGER,WESLEY
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG431502
DRY GAS EXPIRY DATE: 11/11/2016
VIOLATION CODE: DUI
ARREST TIME: 02:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:13

SUBJECT'S LAST NAME: THORPE

DATE/TIME OF TEST: 01/06/2016 02:48:25

Diagnostics Check	OK	02:41
Air Blank	0.000	02:42
Control Test	0.081	02:42
Air Blank	0.000	02:43
Subject Sample # 1	0.159	02:43
Breath Volume	3.281	02:43
Air Blank	0.000	02:44
Air Blank	0.000	02:46
Subject Sample # 2	0.157	02:46
Breath Volume	2.683	02:46
Air Blank	0.000	02:47
Control Test	0.080	02:47
Air Blank	0.000	02:48
Diagnostics Check	OK	02:48

Data Download Date/Time: 01/08/2016 12:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUSSAK,KEVIN,M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: RUSSAK,KEVIN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG431502
DRY GAS EXPIRY DATE: 11/11/2016
VIOLATION CODE: DUI
ARREST TIME: 21:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: KEY

DATE/TIME OF TEST: 01/06/2016 23:02:07

Diagnostics Check	OK	22:55
Air Blank	0.000	22:55
Control Test	0.079	22:56
Air Blank	0.000	22:56
Subject Sample # 1	0.209	22:57
Breath Volume	2.769	22:57
Air Blank	0.000	22:57
Air Blank	0.000	22:59
Subject Sample # 2	0.211	23:00
Breath Volume	2.726	23:00
Air Blank	0.000	23:01
Control Test	0.079	23:01
Air Blank	0.000	23:02
Diagnostics Check	OK	23:02

Data Download Date/Time: 01/08/2016 12:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPENCER, RAYMOND, L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: SPENCER, RAYMOND
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG431502
DRY GAS EXPIRY DATE: 11/11/2016
VIOLATION CODE: DUI
ARREST TIME: 00:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:46

SUBJECT'S LAST NAME: GABBARD

DATE/TIME OF TEST: 01/07/2016 01:14:19

Diagnostics Check	OK	01:06
Air Blank	0.000	01:07
Control Test	0.079	01:07
Air Blank	0.000	01:07
Subject Sample # 1	0.203	01:09
Breath Volume	1.976	01:09
Air Blank	0.000	01:10
Air Blank	0.000	01:12
Subject Sample # 2	0.206	01:12
Breath Volume	1.632	01:12
Air Blank	0.000	01:13
Control Test	0.078	01:13
Air Blank	0.000	01:14
Diagnostics Check	OK	01:14

Data Download Date/Time: 01/08/2016 12:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VITALE,MICHAEL,L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: VITALE,MICHAEL
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 01/08/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG431502
DRY GAS EXPIRY DATE: 11/11/2016
VIOLATION CODE: DUI
ARREST TIME: 20:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:15

SUBJECT'S LAST NAME: BANKICH

DATE/TIME OF TEST: 01/08/2016 21:43:56

Diagnostics Check	OK	21:35	
Air Blank	0.000	21:35	
Control Test	0.079	21:36	
Air Blank	0.000	21:36	
Subject Sample	# 1	21:39	No Sample Provided
Breath Volume	0.000	21:39	
Air Blank	0.000	21:40	
Air Blank	0.000	21:42	
Subject Sample	# 2	21:42	Subject Test Refused
Breath Volume	0.000	21:42	
Air Blank	0.000	21:43	
Control Test	0.078	21:43	
Air Blank	0.000	21:43	
Diagnostics Check	OK	21:43	

Data Download Date/Time: 02/26/2016 16:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBINSON,CHRISTOPHER,A
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: PASCAL,THOMAS
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 01/08/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG431502
DRY GAS EXPIRY DATE: 11/11/2016
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:18

SUBJECT'S LAST NAME: MORRIS

DATE/TIME OF TEST: 01/09/2016 03:56:15

Diagnostics Check	OK	03:49
Air Blank	0.000	03:49
Control Test	0.078	03:49
Air Blank	0.000	03:50
Subject Sample # 1	0.161	03:51
Breath Volume	1.539	03:51
Air Blank	0.000	03:51
Air Blank	0.000	03:53
Subject Sample # 2	0.150	03:54
Breath Volume	2.039	03:54
Air Blank	0.000	03:55
Control Test	0.078	03:55
Air Blank	0.000	03:56
Diagnostics Check	OK	03:56

Data Download Date/Time: 02/26/2016 16:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRUMBIR, EMERALD, G
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: GRUMBIR, EMERALD
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 01/08/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG431502
DRY GAS EXPIRY DATE: 11/11/2016
VIOLATION CODE: DUI
ARREST TIME: 20:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:34

SUBJECT'S LAST NAME: CHRISCO

DATE/TIME OF TEST: 01/14/2016 21:01:34

Diagnostics Check	OK	20:54
Air Blank	0.000	20:55
Control Test	0.080	20:55
Air Blank	0.000	20:55
Subject Sample # 1	0.029	20:57
Breath Volume	2.968	20:57
Air Blank	0.000	20:57
Air Blank	0.000	20:59
Subject Sample # 2	0.027	21:00
Breath Volume	1.726	21:00
Air Blank	0.000	21:00
Control Test	0.078	21:00
Air Blank	0.000	21:01
Diagnostics Check	OK	21:01

Data Download Date/Time: 02/26/2016 16:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERGUSON,DON,L
BREATH TEST OPERATOR AGENCY: Milton PD
ARRESTING OFFICER: ENNIS,SAMUEL
ARRESTING OFFICER AGENCY: Milton PD
LAST AGENCY INSPECTION: 01/08/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG431502
DRY GAS EXPIRY DATE: 11/11/2016
VIOLATION CODE: DUI
ARREST TIME: 00:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 01/15/2016 01:40:13

Diagnostics Check	OK	01:36	
Air Blank	0.000	01:37	
Control Test	0.078	01:37	
Air Blank	0.000	01:37	
Subject Sample	# 1	01:38	Subject Test Refused
Breath Volume	0.000	01:38	
Air Blank	0.000	01:39	
Control Test	0.079	01:39	
Air Blank	0.000	01:40	
Diagnostics Check	OK	01:40	

Data Download Date/Time: 02/26/2016 16:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROWELL,DAVID,M
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: ROWELL,DAVID
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 01/08/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG431502
DRY GAS EXPIRY DATE: 11/11/2016
VIOLATION CODE: DUI
ARREST TIME: 18:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:05

SUBJECT'S LAST NAME: SCHUPP

DATE/TIME OF TEST: 01/15/2016 19:37:42

Diagnostics Check	OK	19:36	
Air Blank		19:37	RFI Detect
Air Blank	0.000	19:37	

Data Download Date/Time: 02/26/2016 16:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROWELL,DAVID,M
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: ROWELL,DAVID
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 01/08/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG431502
DRY GAS EXPIRY DATE: 11/11/2016
VIOLATION CODE: DUI
ARREST TIME: 18:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:05

SUBJECT'S LAST NAME: SCHUPP

DATE/TIME OF TEST: 01/15/2016 19:39:56

Diagnostics Check	OK	19:39	
Air Blank		19:39	RFI Detect
Air Blank		19:39	RFI Detect

Data Download Date/Time: 02/26/2016 16:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROWELL,DAVID,M
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: ROWELL,DAVID
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 01/08/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG431502
DRY GAS EXPIRY DATE: 11/11/2016
VIOLATION CODE: DUI
ARREST TIME: 18:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:05

SUBJECT'S LAST NAME: SCHUPP

DATE/TIME OF TEST: 01/15/2016 19:49:05

Diagnostics Check	OK	19:42
Air Blank	0.000	19:42
Control Test	0.079	19:42
Air Blank	0.000	19:43
Subject Sample # 1	0.092	19:44
Breath Volume	1.187	19:44
Air Blank	0.000	19:45
Air Blank	0.000	19:47
Subject Sample # 2	0.095	19:47
Breath Volume	2.183	19:47
Air Blank	0.000	19:48
Control Test	0.079	19:48
Air Blank	0.000	19:48
Diagnostics Check	OK	19:49

Data Download Date/Time: 02/26/2016 16:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PASCHAL,THOMAS,C
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: PASCHAL,THOMAS
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 01/08/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG431502
DRY GAS EXPIRY DATE: 11/11/2016
VIOLATION CODE: DUI
ARREST TIME: 01:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:11

SUBJECT'S LAST NAME: PETERSON

DATE/TIME OF TEST: 01/16/2016 03:02:15

Diagnostics Check	OK	02:54
Air Blank	0.000	02:55
Control Test	0.079	02:55
Air Blank	0.000	02:56
Subject Sample # 1	0.230	02:57
Breath Volume	2.093	02:57
Air Blank	0.000	02:58
Air Blank	0.000	03:00
Subject Sample # 2	0.220	03:00
Breath Volume	1.535	03:00
Air Blank	0.000	03:01
Control Test	0.078	03:01
Air Blank	0.000	03:02
Diagnostics Check	OK	03:02

Data Download Date/Time: 02/26/2016 16:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VITALE,MICHAEL,L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: VITALE,MICHAEL
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 01/08/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG431502
DRY GAS EXPIRY DATE: 11/11/2016
VIOLATION CODE: DUI
ARREST TIME: 17:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:44

SUBJECT'S LAST NAME: BYTELL

DATE/TIME OF TEST: 01/16/2016 19:07:34

Diagnostics Check	OK	19:04	
Air Blank	0.000	19:05	
Control Test	0.079	19:05	
Air Blank	0.000	19:06	
Subject Sample	# 1	19:06	Subject Test Refused
Breath Volume	0.000	19:06	
Air Blank	0.000	19:06	
Control Test	0.079	19:06	
Air Blank	0.000	19:07	
Diagnostics Check	OK	19:07	

Data Download Date/Time: 02/26/2016 16:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ENFINGER, THOMAS, D
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: ENFINGER, THOMAS
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 01/08/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG431502
DRY GAS EXPIRY DATE: 11/11/2016
VIOLATION CODE: DUI
ARREST TIME: 23:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: GRAY

DATE/TIME OF TEST: 01/17/2016 00:23:56

Diagnostics Check	OK	00:21	
Air Blank	0.000	00:21	
Control Test	0.080	00:21	
Air Blank	0.000	00:22	
Subject Sample	# 1	00:22	Subject Test Refused
Breath Volume	0.000	00:22	
Air Blank	0.000	00:23	
Control Test	0.079	00:23	
Air Blank	0.000	00:23	
Diagnostics Check	OK	00:23	

Data Download Date/Time: 02/26/2016 16:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLEMENTS, WILLIAM, J
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: CLEMENTS, WILLIAM
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 01/08/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG431502
DRY GAS EXPIRY DATE: 11/11/2016
VIOLATION CODE: DUI
ARREST TIME: 01:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:24

SUBJECT'S LAST NAME: DERENBECKER

DATE/TIME OF TEST: 01/17/2016 02:53:13

Diagnostics Check	OK	02:44	
Air Blank	0.000	02:44	
Control Test	0.079	02:45	
Air Blank	0.000	02:45	
Subject Sample	# 1	02:48	No Sample Provided
Breath Volume	0.000	02:48	
Air Blank	0.000	02:49	
Air Blank	0.000	02:51	
Subject Sample	# 2	02:51	Subject Test Refused
Breath Volume	0.000	02:51	
Air Blank	0.000	02:52	
Control Test	0.080	02:52	
Air Blank	0.000	02:53	
Diagnostics Check	OK	02:53	

Data Download Date/Time: 02/26/2016 16:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRIBB,BRIAN,S
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: AGUIAR,KIMBERLY
ARRESTING OFFICER AGENCY: Milton PD
LAST AGENCY INSPECTION: 01/08/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG431502
DRY GAS EXPIRY DATE: 11/11/2016
VIOLATION CODE: DUI
ARREST TIME: 17:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:33

SUBJECT'S LAST NAME: HILLIARD

DATE/TIME OF TEST: 01/18/2016 18:57:01

Diagnostics Check	OK	18:53	
Air Blank	0.000	18:54	
Control Test	0.078	18:54	
Air Blank	0.000	18:54	
Subject Sample	# 1	18:56	Sequence Aborted
Breath Volume	0.000	18:56	
Air Blank	0.000	18:57	

Data Download Date/Time: 02/26/2016 16:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPENCER, RAYMOND, L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: SPENCER, RAYMOND
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 01/08/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG431502
DRY GAS EXPIRY DATE: 11/11/2016
VIOLATION CODE: DUI
ARREST TIME: 23:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 01/24/2016 00:32:47

Diagnostics Check	OK	00:25
Air Blank	0.000	00:26
Control Test	0.080	00:26
Air Blank	0.000	00:27
Subject Sample # 1	0.000	00:28
Breath Volume	2.597	00:28
Air Blank	0.000	00:28
Air Blank	0.000	00:30
Subject Sample # 2	0.000	00:31
Breath Volume	1.988	00:31
Air Blank	0.000	00:31
Control Test	0.080	00:32
Air Blank	0.000	00:32
Diagnostics Check	OK	00:32

Data Download Date/Time: 02/26/2016 16:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAHL,RANDY,J
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: DAHL,RANDY
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 01/08/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG431502
DRY GAS EXPIRY DATE: 11/11/2016
VIOLATION CODE: DUI
ARREST TIME: 01:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: BOWMAN

DATE/TIME OF TEST: 01/24/2016 02:09:38

Diagnostics Check	OK	02:03
Air Blank	0.000	02:03
Control Test	0.081	02:04
Air Blank	0.000	02:04
Subject Sample # 1	0.000	02:05
Breath Volume	2.195	02:05
Air Blank	0.000	02:05
Air Blank	0.000	02:07
Subject Sample # 2	0.000	02:08
Breath Volume	2.000	02:08
Air Blank	0.000	02:08
Control Test	0.079	02:09
Air Blank	0.000	02:09
Diagnostics Check	OK	02:09

Data Download Date/Time: 02/26/2016 16:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUSSAK,KEVIN,M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: RUSSAK,KEVIN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 01/08/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG431502
DRY GAS EXPIRY DATE: 11/11/2016
VIOLATION CODE: DUI
ARREST TIME: 21:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: GRAHAM

DATE/TIME OF TEST: 01/26/2016 22:54:49

Diagnostics Check	OK	22:47
Air Blank	0.000	22:47
Control Test	0.079	22:48
Air Blank	0.000	22:48
Subject Sample # 1	0.245	22:50
Breath Volume	3.343	22:50
Air Blank	0.000	22:50
Air Blank	0.000	22:52
Subject Sample # 2	0.238	22:53
Breath Volume	4.343	22:53
Air Blank	0.000	22:53
Control Test	0.078	22:54
Air Blank	0.000	22:54
Diagnostics Check	OK	22:54

Data Download Date/Time: 02/26/2016 16:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALLACE,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: WALLACE,CHRISTOPHER
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 01/08/2016

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: AG431502
DRY GAS EXPIRY DATE: 11/11/2016
VIOLATION CODE: DUI
ARREST TIME: 22:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:05

SUBJECT'S LAST NAME: STROTT

DATE/TIME OF TEST: 01/26/2016 23:48:58

Diagnostics Check	OK	23:42
Air Blank	0.000	23:43
Control Test	0.080	23:43
Air Blank	0.000	23:44
Subject Sample # 1	0.000	23:44
Breath Volume	2.578	23:44
Air Blank	0.000	23:45
Air Blank	0.000	23:47
Subject Sample # 2	0.000	23:47
Breath Volume	2.214	23:47
Air Blank	0.000	23:48
Control Test	0.078	23:48
Air Blank	0.000	23:48
Diagnostics Check	OK	23:48

Data Download Date/Time: 02/26/2016 16:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CALABRO,ALFRED,J
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: CALABRO,ALFRED
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 01/08/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG431502
DRY GAS EXPIRY DATE: 11/11/2016
VIOLATION CODE: DUI
ARREST TIME: 01:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: MCCORMICK

DATE/TIME OF TEST: 01/28/2016 02:34:19

Diagnostics Check	OK	02:22	
Air Blank	0.000	02:23	
Control Test	0.078	02:23	
Air Blank	0.000	02:23	
Subject Sample	# 1 0.072	02:27	Volume Not Met
Breath Volume	0.261	02:27	
Air Blank	0.000	02:27	
Air Blank	0.000	02:29	
Subject Sample	# 2 0.067	02:32	Volume Not Met
Breath Volume	0.835	02:32	
Air Blank	0.000	02:33	
Control Test	0.078	02:33	
Air Blank	0.000	02:34	
Diagnostics Check	OK	02:34	

Data Download Date/Time: 02/26/2016 16:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AQUINO,OMAR,E
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: AQUINO,OMAR
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 01/08/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG431502
DRY GAS EXPIRY DATE: 11/11/2016
VIOLATION CODE: DUI
ARREST TIME: 15:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:54

SUBJECT'S LAST NAME: BLACKWELL

DATE/TIME OF TEST: 01/28/2016 16:49:19

Diagnostics Check	OK	16:42
Air Blank	0.000	16:43
Control Test	0.078	16:43
Air Blank	0.000	16:44
Subject Sample # 1	0.267	16:44
Breath Volume	3.484	16:44
Air Blank	0.000	16:45
Air Blank	0.000	16:47
Subject Sample # 2	0.264	16:47
Breath Volume	3.144	16:47
Air Blank	0.000	16:48
Control Test	0.078	16:48
Air Blank	0.000	16:49
Diagnostics Check	OK	16:49

Data Download Date/Time: 02/26/2016 16:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUSSAK,KEVIN,M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: RUSSAK,KEVIN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 01/08/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG431502
DRY GAS EXPIRY DATE: 11/11/2016
VIOLATION CODE: DUI
ARREST TIME: 18:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:20

SUBJECT'S LAST NAME: CRAIN

DATE/TIME OF TEST: 01/28/2016 19:29:07

Diagnostics Check	OK	19:22
Air Blank	0.000	19:22
Control Test	0.079	19:23
Air Blank	0.000	19:23
Subject Sample # 1	0.222	19:24
Breath Volume	2.941	19:24
Air Blank	0.000	19:25
Air Blank	0.000	19:27
Subject Sample # 2	0.218	19:27
Breath Volume	2.800	19:27
Air Blank	0.000	19:28
Control Test	0.080	19:28
Air Blank	0.000	19:29
Diagnostics Check	OK	19:29

Data Download Date/Time: 02/26/2016 16:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SWINDELL,SHAWN,T
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: SWINDELL,SHAWN
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 01/08/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG431502
DRY GAS EXPIRY DATE: 11/11/2016
VIOLATION CODE: DUI
ARREST TIME: 23:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:57

SUBJECT'S LAST NAME: MCCAULEY

DATE/TIME OF TEST: 01/30/2016 00:23:47

Diagnostics Check	OK	00:17
Air Blank	0.000	00:17
Control Test	0.078	00:18
Air Blank	0.000	00:18
Subject Sample # 1	0.146	00:19
Breath Volume	2.535	00:19
Air Blank	0.000	00:19
Air Blank	0.000	00:21
Subject Sample # 2	0.152	00:22
Breath Volume	2.746	00:22
Air Blank	0.000	00:22
Control Test	0.077	00:23
Air Blank	0.000	00:23
Diagnostics Check	OK	00:23

Data Download Date/Time: 02/26/2016 16:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLARK,STEPHEN,F
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: CLARK,STEPHEN
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 01/08/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG431502
DRY GAS EXPIRY DATE: 11/11/2016
VIOLATION CODE: DUI
ARREST TIME: 17:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:00

SUBJECT'S LAST NAME: BUECHNER

DATE/TIME OF TEST: 01/30/2016 18:29:42

Diagnostics Check	OK	18:21	
Air Blank	0.000	18:21	
Control Test	0.079	18:21	
Air Blank	0.000	18:22	
Subject Sample	# 1	18:25	No Sample Provided
Breath Volume	0.000	18:25	
Air Blank	0.000	18:25	
Air Blank	0.000	18:27	
Subject Sample	# 2	18:28	Subject Test Refused
Breath Volume	0.000	18:28	
Air Blank	0.000	18:28	
Control Test	0.079	18:29	
Air Blank	0.000	18:29	
Diagnostics Check	OK	18:29	

Data Download Date/Time: 02/26/2016 16:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHAU, LIEM, V
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: CHAU, LIEM
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 01/08/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG431502
DRY GAS EXPIRY DATE: 11/11/2016
VIOLATION CODE: DUI
ARREST TIME: 20:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:21

SUBJECT'S LAST NAME: CURTIS

DATE/TIME OF TEST: 01/30/2016 21:47:36

Diagnostics Check	OK	21:41
Air Blank	0.000	21:41
Control Test	0.079	21:42
Air Blank	0.000	21:42
Subject Sample # 1	0.319	21:43
Breath Volume	2.816	21:43
Air Blank	0.000	21:43
Air Blank	0.000	21:45
Subject Sample # 2	0.317	21:46
Breath Volume	3.746	21:46
Air Blank	0.000	21:46
Control Test	0.078	21:47
Air Blank	0.000	21:47
Diagnostics Check	OK	21:47

Data Download Date/Time: 02/26/2016 16:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIGHTOWER,CONNIE,A
BREATH TEST OPERATOR AGENCY: Lake City PD
ARRESTING OFFICER: GESICKI,L
ARRESTING OFFICER AGENCY: Lake City PD
LAST AGENCY INSPECTION: 12/28/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1686772
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 06:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:27

SUBJECT'S LAST NAME: WADE

DATE/TIME OF TEST: 01/01/2016 06:53:56

Diagnostics Check	OK	06:47
Air Blank	0.000	06:47
Control Test	0.080	06:48
Air Blank	0.000	06:48
Subject Sample # 1	0.278	06:49
Breath Volume	3.527	06:49
Air Blank	0.000	06:50
Air Blank	0.000	06:51
Subject Sample # 2	0.277	06:52
Breath Volume	3.207	06:52
Air Blank	0.000	06:53
Control Test	0.078	06:53
Air Blank	0.000	06:53
Diagnostics Check	OK	06:53

Data Download Date/Time: 01/24/2016 15:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,CARLOS,R
BREATH TEST OPERATOR AGENCY: Lake City PD
ARRESTING OFFICER: MAYS,CHAUNCEY
ARRESTING OFFICER AGENCY: Lake City PD
LAST AGENCY INSPECTION: 12/28/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1686772
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 19:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:20

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 01/05/2016 20:56:13

Diagnostics Check	OK	20:49
Air Blank	0.000	20:50
Control Test	0.080	20:50
Air Blank	0.000	20:50
Subject Sample # 1	0.236	20:51
Breath Volume	1.250	20:51
Air Blank	0.000	20:52
Air Blank	0.000	20:54
Subject Sample # 2	0.231	20:54
Breath Volume	1.726	20:54
Air Blank	0.000	20:55
Control Test	0.078	20:55
Air Blank	0.000	20:56
Diagnostics Check	OK	20:56

Data Download Date/Time: 01/24/2016 15:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLIE,JOHN,R
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: TILLIE,JOHN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 12/28/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1686772
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 01:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: BROWNING J

DATE/TIME OF TEST: 01/07/2016 03:16:13

Diagnostics Check	OK	03:13	
Air Blank	0.000	03:13	
Control Test	0.081	03:14	
Air Blank	0.000	03:14	
Subject Sample	# 1	03:14	Subject Test Refused
Breath Volume	0.000	03:14	
Air Blank	0.000	03:15	
Control Test	0.081	03:15	
Air Blank	0.000	03:16	
Diagnostics Check	OK	03:16	

Data Download Date/Time: 01/24/2016 15:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALKER,SHELIA,
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: WALKER,SHELIA
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 12/28/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1686772
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 14:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:04

SUBJECT'S LAST NAME: VEGA JUSTINIANO

DATE/TIME OF TEST: 01/10/2016 03:32:16

Diagnostics Check	OK	03:24
Air Blank	0.000	03:25
Control Test	0.080	03:25
Air Blank	0.000	03:25
Subject Sample # 1	0.081	03:27
Breath Volume	1.796	03:27
Air Blank	0.000	03:28
Air Blank	0.000	03:29
Subject Sample # 2	0.080	03:30
Breath Volume	1.753	03:30
Air Blank	0.000	03:31
Control Test	0.078	03:31
Air Blank	0.000	03:32
Diagnostics Check	OK	03:32

Data Download Date/Time: 01/24/2016 15:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MACLAREN,DEREK,L
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: TILLIE,JOHN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 12/28/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1686772
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 12:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:22

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 01/13/2016 01:48:43

Diagnostics Check	OK	01:42
Air Blank	0.000	01:42
Control Test	0.080	01:42
Air Blank	0.000	01:43
Subject Sample # 1	0.135	01:44
Breath Volume	3.058	01:44
Air Blank	0.000	01:44
Air Blank	0.000	01:46
Subject Sample # 2	0.125	01:47
Breath Volume	1.710	01:47
Air Blank	0.000	01:47
Control Test	0.080	01:48
Air Blank	0.000	01:48
Diagnostics Check	OK	01:48

Data Download Date/Time: 01/24/2016 15:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000780

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WADFORD,CHARLES,T
BREATH TEST OPERATOR AGENCY: Suwannee County SO
ARRESTING OFFICER: WADFORD,CHARLES
ARRESTING OFFICER AGENCY: Suwannee County SO
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06415080A1
DRY GAS EXPIRY DATE: 04/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: WELLS

DATE/TIME OF TEST: 01/12/2016 04:10:00

Diagnostics Check	OK	04:04	
Air Blank	0.000	04:05	
Control Test	0.083	04:05	
Air Blank	0.000	04:05	
Subject Sample	# 1	04:08	Subject Test Refused
Breath Volume	0.000	04:08	
Air Blank	0.000	04:09	
Control Test	0.084	04:09	
Air Blank	0.000	04:09	
Diagnostics Check	OK	04:10	

Data Download Date/Time: 01/31/2016 05:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000780

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAYLOR,JAMES,R
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: TAYLOR,JAMES
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06415080A1
DRY GAS EXPIRY DATE: 04/05/2017
VIOLATION CODE: DUI
ARREST TIME: 07:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:33

SUBJECT'S LAST NAME: BOATRIGHT

DATE/TIME OF TEST: 01/24/2016 08:57:44

Diagnostics Check	OK	08:54	
Air Blank	0.000	08:55	
Control Test	0.082	08:55	
Air Blank	0.000	08:56	
Subject Sample	# 1	08:56	RFI Detect
Breath Volume	0.000	08:56	
Air Blank	0.000	08:56	
Control Test	0.083	08:57	
Air Blank	0.000	08:57	
Diagnostics Check	OK	08:57	

Data Download Date/Time: 01/31/2016 05:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000780

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAYLOR,JAMES,R
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: TAYLOR,JAMES
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06415080A1
DRY GAS EXPIRY DATE: 04/05/2017
VIOLATION CODE: DUI
ARREST TIME: 07:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:33

SUBJECT'S LAST NAME: BOATRIGHT

DATE/TIME OF TEST: 01/24/2016 09:02:57

Diagnostics Check	OK	08:59	
Air Blank	0.000	09:00	
Control Test	0.083	09:00	
Air Blank	0.000	09:01	
Subject Sample	# 1 0.194	09:01	
Breath Volume	3.726	09:01	
Air Blank		09:02	Purge Fail
Air Blank	0.000	09:02	

Data Download Date/Time: 01/31/2016 05:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000780

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAYLOR,JAMES,R
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: TAYLOR,JAMES
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06415080A1
DRY GAS EXPIRY DATE: 04/05/2017
VIOLATION CODE: DUI
ARREST TIME: 07:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:33

SUBJECT'S LAST NAME: BOATRIGHT

DATE/TIME OF TEST: 01/24/2016 09:11:37

Diagnostics Check	OK	09:04
Air Blank	0.000	09:05
Control Test	0.080	09:05
Air Blank	0.000	09:06
Subject Sample # 1	0.187	09:07
Breath Volume	3.535	09:07
Air Blank	0.000	09:08
Air Blank	0.000	09:09
Subject Sample # 2	0.184	09:09
Breath Volume	2.148	09:09
Air Blank	0.000	09:10
Control Test	0.081	09:11
Air Blank	0.000	09:11
Diagnostics Check	OK	09:11

Data Download Date/Time: 01/31/2016 05:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000780

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TOMPKINS,CHARLES,P
BREATH TEST OPERATOR AGENCY: Suwannee County SO
ARRESTING OFFICER: TOMPKINS,CHARLES
ARRESTING OFFICER AGENCY: Suwannee County SO
LAST AGENCY INSPECTION: 01/31/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06415080A1
DRY GAS EXPIRY DATE: 04/05/2017
VIOLATION CODE: DUI
ARREST TIME: 18:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:20

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 01/31/2016 19:58:04

Diagnostics Check	OK	19:49
Air Blank	0.000	19:49
Control Test	0.083	19:50
Air Blank	0.000	19:50
Subject Sample # 1	0.185	19:52
Breath Volume	1.824	19:52
Air Blank	0.000	19:53
Air Blank	0.000	19:54
Subject Sample # 2	0.174	19:56
Breath Volume	1.531	19:56
Air Blank	0.000	19:57
Control Test	0.083	19:57
Air Blank	0.000	19:57
Diagnostics Check	OK	19:58

Data Download Date/Time: 02/29/2016 02:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000785

INSTRUMENT REGISTERED TO: HOLMES COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PATE,THOMAS,W
BREATH TEST OPERATOR AGENCY: Holmes County SO
ARRESTING OFFICER: PATE,THOMAS
ARRESTING OFFICER AGENCY: Holmes County SO
LAST AGENCY INSPECTION: 12/01/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20314080A4
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 22:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: STEWART

DATE/TIME OF TEST: 01/02/2016 23:52:32

Diagnostics Check	OK	23:45
Air Blank	0.000	23:45
Control Test	0.079	23:46
Air Blank	0.000	23:46
Subject Sample # 1	0.162	23:47
Breath Volume	2.328	23:47
Air Blank	0.000	23:48
Air Blank	0.000	23:50
Subject Sample # 2	0.167	23:50
Breath Volume	1.441	23:50
Air Blank	0.000	23:51
Control Test	0.078	23:51
Air Blank	0.000	23:52
Diagnostics Check	OK	23:52

Data Download Date/Time: 03/05/2016 17:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000785

INSTRUMENT REGISTERED TO: HOLMES COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VENNELL,JOHN,B
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: VENNELL,JOHN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 12/01/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20314080A4
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: HENDRIX

DATE/TIME OF TEST: 01/03/2016 00:30:17

Diagnostics Check	OK	00:24
Air Blank	0.000	00:24
Control Test	0.080	00:24
Air Blank	0.000	00:25
Subject Sample # 1	0.136	00:25
Breath Volume	2.789	00:25
Air Blank	0.000	00:26
Air Blank	0.000	00:28
Subject Sample # 2	0.134	00:28
Breath Volume	2.812	00:28
Air Blank	0.000	00:29
Control Test	0.078	00:29
Air Blank	0.000	00:30
Diagnostics Check	OK	00:30

Data Download Date/Time: 03/05/2016 17:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000785

INSTRUMENT REGISTERED TO: HOLMES COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOWMAN,JOANNA,N
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BOWMAN,JOANNA
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 12/01/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20314080A4
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 20:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: PATTERSON

DATE/TIME OF TEST: 01/03/2016 22:04:26

Diagnostics Check	OK	21:57
Air Blank	0.000	21:58
Control Test	0.079	21:58
Air Blank	0.000	21:58
Subject Sample # 1	0.149	22:00
Breath Volume	1.906	22:00
Air Blank	0.000	22:00
Air Blank	0.000	22:02
Subject Sample # 2	0.143	22:02
Breath Volume	1.660	22:02
Air Blank	0.000	22:03
Control Test	0.079	22:03
Air Blank	0.000	22:04
Diagnostics Check	OK	22:04

Data Download Date/Time: 03/05/2016 17:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000785

INSTRUMENT REGISTERED TO: HOLMES COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARVER,JASION,G
BREATH TEST OPERATOR AGENCY: Bonifay PD
ARRESTING OFFICER: CARVER,JASION
ARRESTING OFFICER AGENCY: Bonifay PD
LAST AGENCY INSPECTION: 12/01/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20314080A4
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 23:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:08

SUBJECT'S LAST NAME: BEAVER

DATE/TIME OF TEST: 01/05/2016 00:44:06

Diagnostics Check	OK	00:29	
Air Blank	0.000	00:30	
Control Test	0.079	00:30	
Air Blank	0.000	00:30	
Subject Sample	# 1 0.183	00:34	Volume Not Met
Breath Volume	0.957	00:34	
Air Blank	0.000	00:34	
Air Blank	0.000	00:36	
Subject Sample	# 2 0.206	00:38	
Breath Volume	1.484	00:38	
Air Blank	0.000	00:39	
Air Blank	0.000	00:41	
Subject Sample	# 3 0.196	00:42	
Breath Volume	1.261	00:42	
Air Blank	0.000	00:43	
Control Test	0.076	00:43	
Air Blank	0.000	00:44	
Diagnostics Check	OK	00:44	

Data Download Date/Time: 03/05/2016 17:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000785

INSTRUMENT REGISTERED TO: HOLMES COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DORIS,BRANDI,M
BREATH TEST OPERATOR AGENCY: Bonifay PD
ARRESTING OFFICER: LONG,J
ARRESTING OFFICER AGENCY: Bonifay PD
LAST AGENCY INSPECTION: 12/01/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20314080A4
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 16:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:05

SUBJECT'S LAST NAME: GRAY

DATE/TIME OF TEST: 01/28/2016 17:31:59

Diagnostics Check	OK	17:25
Air Blank	0.000	17:26
Control Test	0.079	17:26
Air Blank	0.000	17:26
Subject Sample # 1	0.306	17:27
Breath Volume	2.050	17:27
Air Blank	0.000	17:28
Air Blank	0.000	17:29
Subject Sample # 2	0.318	17:30
Breath Volume	1.835	17:30
Air Blank	0.000	17:31
Control Test	0.078	17:31
Air Blank	0.000	17:31
Diagnostics Check	OK	17:31

Data Download Date/Time: 03/05/2016 17:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000785

INSTRUMENT REGISTERED TO: HOLMES COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BASS,JAMES,G
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BASS,JAMES
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 12/01/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20314080A4
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 15:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:00

SUBJECT'S LAST NAME: MCDONALD

DATE/TIME OF TEST: 01/30/2016 18:14:47

Diagnostics Check	OK	18:08
Air Blank	0.000	18:08
Control Test	0.078	18:08
Air Blank	0.000	18:09
Subject Sample # 1	0.285	18:10
Breath Volume	1.710	18:10
Air Blank	0.000	18:11
Air Blank	0.000	18:12
Subject Sample # 2	0.292	18:13
Breath Volume	2.335	18:13
Air Blank	0.000	18:13
Control Test	0.078	18:14
Air Blank	0.000	18:14
Diagnostics Check	OK	18:14

Data Download Date/Time: 03/05/2016 17:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000786

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARNANDAN, INDROWTIE,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SERPHOS, STEVEN
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 01/01/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 20:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:20

SUBJECT'S LAST NAME: JAROSZEWSKI

DATE/TIME OF TEST: 01/01/2016 21:09:58

Diagnostics Check	OK	21:03
Air Blank	0.000	21:04
Control Test	0.079	21:04
Air Blank	0.000	21:05
Subject Sample # 1	0.134	21:05
Breath Volume	1.937	21:05
Air Blank	0.000	21:06
Air Blank	0.000	21:08
Subject Sample # 2	0.134	21:08
Breath Volume	1.699	21:08
Air Blank	0.000	21:09
Control Test	0.079	21:09
Air Blank	0.000	21:09
Diagnostics Check	OK	21:09

Data Download Date/Time: 01/01/2016 22:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERSON,THOMAS,C
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: ROBERSON,THOMAS
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 12/15/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 17:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:10

SUBJECT'S LAST NAME: CHAVERS

DATE/TIME OF TEST: 01/04/2016 18:44:39

Diagnostics Check	OK	18:39	
Air Blank	0.000	18:40	
Control Test	0.080	18:40	
Air Blank	0.000	18:40	
Subject Sample	# 1	18:43	Interferent Detect
Breath Volume	1.972	18:43	
Air Blank		18:44	RFI Detect
Air Blank		18:44	RFI Detect

Data Download Date/Time: 01/23/2016 11:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TATRO,KERSTAN,J
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: TATRO,KERSTAN
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 12/15/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:19

SUBJECT'S LAST NAME: SUMMERS

DATE/TIME OF TEST: 01/06/2016 02:42:29

Diagnostics Check	OK	02:39	
Air Blank	0.000	02:39	
Control Test	0.079	02:40	
Air Blank	0.000	02:40	
Subject Sample	# 1	02:41	Subject Test Refused
Breath Volume	0.000	02:41	
Air Blank	0.000	02:41	
Control Test	0.081	02:41	
Air Blank	0.000	02:42	
Diagnostics Check	OK	02:42	

Data Download Date/Time: 01/23/2016 11:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TATRO,KERSTAN,J
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: TATRO,KERSTAN
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 12/15/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:37

SUBJECT'S LAST NAME: BURGESS

DATE/TIME OF TEST: 01/10/2016 02:03:26

Diagnostics Check	OK	02:00	
Air Blank	0.000	02:01	
Control Test	0.079	02:01	
Air Blank	0.000	02:01	
Subject Sample	# 1	02:02	Subject Test Refused
Breath Volume	0.000	02:02	
Air Blank	0.000	02:02	
Control Test	0.079	02:02	
Air Blank	0.000	02:03	
Diagnostics Check	OK	02:03	

Data Download Date/Time: 01/23/2016 11:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITE,TRAVIS,D
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: KIDD,JERMEL
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 12/15/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: KRAUSE

DATE/TIME OF TEST: 01/18/2016 00:52:20

Diagnostics Check	OK	00:46
Air Blank	0.000	00:46
Control Test	0.079	00:46
Air Blank	0.000	00:47
Subject Sample # 1	0.122	00:47
Breath Volume	1.210	00:47
Air Blank	0.000	00:48
Air Blank	0.000	00:50
Subject Sample # 2	0.129	00:50
Breath Volume	1.312	00:50
Air Blank	0.000	00:51
Control Test	0.080	00:51
Air Blank	0.000	00:52
Diagnostics Check	OK	00:52

Data Download Date/Time: 01/23/2016 11:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TANTAY,R,D
BREATH TEST OPERATOR AGENCY: OTHER Fed
ARRESTING OFFICER: TANTAY,R
ARRESTING OFFICER AGENCY: OTHER Fed
LAST AGENCY INSPECTION: 01/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:55

SUBJECT'S LAST NAME: SWANN

DATE/TIME OF TEST: 01/23/2016 21:27:25

Diagnostics Check	OK	21:16	
Air Blank	0.000	21:16	
Control Test	0.080	21:16	
Air Blank	0.000	21:17	
Subject Sample	# 1	21:20	No Sample Provided
Breath Volume	0.000	21:20	
Air Blank	0.000	21:20	
Air Blank	0.000	21:22	
Subject Sample	# 2	21:26	No Sample Provided
Breath Volume	0.000	21:26	
Air Blank	0.000	21:26	
Control Test	0.080	21:26	
Air Blank	0.000	21:27	
Diagnostics Check	OK	21:27	

Data Download Date/Time: 02/15/2016 10:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCAPIN,BRANDON,L
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: SCAPIN,BRANDON
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 01/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:53

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 01/28/2016 00:16:48

Diagnostics Check	OK	00:13	
Air Blank	0.000	00:14	
Control Test	0.079	00:14	
Air Blank	0.000	00:14	
Subject Sample	# 1	00:15	Subject Test Refused
Breath Volume	0.000	00:15	
Air Blank	0.000	00:15	
Control Test	0.079	00:16	
Air Blank	0.000	00:16	
Diagnostics Check	OK	00:16	

Data Download Date/Time: 02/15/2016 10:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOGAN,TODD,A
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: HOGAN,TODD
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 12/29/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26614080A7
DRY GAS EXPIRY DATE: 10/05/2016
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:52

SUBJECT'S LAST NAME: ALBERT

DATE/TIME OF TEST: 01/02/2016 03:20:29

Diagnostics Check	OK	03:14
Air Blank	0.000	03:14
Control Test	0.080	03:15
Air Blank	0.000	03:15
Subject Sample # 1	0.137	03:16
Breath Volume	3.113	03:16
Air Blank	0.000	03:16
Air Blank	0.000	03:18
Subject Sample # 2	0.133	03:18
Breath Volume	2.035	03:18
Air Blank	0.000	03:19
Control Test	0.080	03:19
Air Blank	0.000	03:20
Diagnostics Check	OK	03:20

Data Download Date/Time: 01/12/2016 14:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASSERLY,STEPHEN,R
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: CASSERLY,STEPHEN
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 12/29/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26614080A7
DRY GAS EXPIRY DATE: 10/05/2016
VIOLATION CODE: DUI
ARREST TIME: 02:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:06

SUBJECT'S LAST NAME: WEST

DATE/TIME OF TEST: 01/02/2016 03:50:24

Diagnostics Check	OK	03:43
Air Blank	0.000	03:43
Control Test	0.079	03:44
Air Blank	0.000	03:44
Subject Sample # 1	0.000	03:46
Breath Volume	2.734	03:46
Air Blank	0.000	03:46
Air Blank	0.000	03:48
Subject Sample # 2	0.000	03:49
Breath Volume	2.593	03:49
Air Blank	0.000	03:49
Control Test	0.078	03:49
Air Blank	0.000	03:50
Diagnostics Check	OK	03:50

Data Download Date/Time: 01/12/2016 14:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARNER,TODD,A
BREATH TEST OPERATOR AGENCY: Fort Pierce PD
ARRESTING OFFICER: WARNER,TODD
ARRESTING OFFICER AGENCY: Fort Pierce PD
LAST AGENCY INSPECTION: 12/29/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26614080A7
DRY GAS EXPIRY DATE: 10/05/2016
VIOLATION CODE: DUI
ARREST TIME: 23:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:57

SUBJECT'S LAST NAME: MAYO MALDONADO

DATE/TIME OF TEST: 01/06/2016 00:24:23

Diagnostics Check	OK	00:17
Air Blank	0.000	00:18
Control Test	0.081	00:18
Air Blank	0.000	00:19
Subject Sample # 1	0.234	00:19
Breath Volume	3.156	00:19
Air Blank	0.000	00:20
Air Blank	0.000	00:22
Subject Sample # 2	0.234	00:22
Breath Volume	3.386	00:22
Air Blank	0.000	00:23
Control Test	0.080	00:23
Air Blank	0.000	00:24
Diagnostics Check	OK	00:24

Data Download Date/Time: 01/12/2016 14:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOGAN,TODD,A
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: HOGAN,TODD
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 12/29/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26614080A7
DRY GAS EXPIRY DATE: 10/05/2016
VIOLATION CODE: DUI
ARREST TIME: 21:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:06

SUBJECT'S LAST NAME: WOOD

DATE/TIME OF TEST: 01/07/2016 22:35:11

Diagnostics Check	OK	22:29
Air Blank	0.000	22:29
Control Test	0.080	22:29
Air Blank	0.000	22:30
Subject Sample # 1	0.207	22:30
Breath Volume	1.753	22:30
Air Blank	0.000	22:31
Air Blank	0.000	22:33
Subject Sample # 2	0.204	22:33
Breath Volume	1.644	22:33
Air Blank	0.000	22:34
Control Test	0.080	22:34
Air Blank	0.000	22:35
Diagnostics Check	OK	22:35

Data Download Date/Time: 01/12/2016 14:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOGAN,TODD,A
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: HOGAN,TODD
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 12/29/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26614080A7
DRY GAS EXPIRY DATE: 10/05/2016
VIOLATION CODE: DUI
ARREST TIME: 02:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:27

SUBJECT'S LAST NAME: TIFFANY

DATE/TIME OF TEST: 01/10/2016 03:00:27

Diagnostics Check	OK	02:49	
Air Blank	0.000	02:49	
Control Test	0.080	02:49	
Air Blank	0.000	02:50	
Subject Sample	# 1	02:53	No Sample Provided
Breath Volume	0.000	02:53	
Air Blank	0.000	02:53	
Air Blank	0.000	02:55	
Subject Sample	# 2	02:59	No Sample Provided
Breath Volume	0.000	02:59	
Air Blank	0.000	02:59	
Control Test	0.080	02:59	
Air Blank	0.000	03:00	
Diagnostics Check	OK	03:00	

Data Download Date/Time: 01/12/2016 14:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAYFORD,GREGG,R
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: HAYFORD,GREGG
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 01/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26614080A7
DRY GAS EXPIRY DATE: 10/05/2016
VIOLATION CODE: DUI
ARREST TIME: 08:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:48

SUBJECT'S LAST NAME: PIERRE

DATE/TIME OF TEST: 01/16/2016 09:36:35

Diagnostics Check	OK	09:30
Air Blank	0.000	09:30
Control Test	0.080	09:31
Air Blank	0.000	09:31
Subject Sample # 1	0.050	09:32
Breath Volume	1.375	09:32
Air Blank	0.000	09:32
Air Blank	0.000	09:34
Subject Sample # 2	0.048	09:35
Breath Volume	1.500	09:35
Air Blank	0.000	09:35
Control Test	0.081	09:36
Air Blank	0.000	09:36
Diagnostics Check	OK	09:36

Data Download Date/Time: 02/25/2016 12:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FELIX,GEORGE,D
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: FELIX,GEORGE
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 01/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26614080A7
DRY GAS EXPIRY DATE: 10/05/2016
VIOLATION CODE: DUI
ARREST TIME: 00:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: REEDER

DATE/TIME OF TEST: 01/17/2016 01:26:45

Diagnostics Check	OK	01:20
Air Blank	0.000	01:20
Control Test	0.079	01:21
Air Blank	0.000	01:21
Subject Sample # 1	0.213	01:22
Breath Volume	5.148	01:22
Air Blank	0.000	01:22
Air Blank	0.000	01:24
Subject Sample # 2	0.214	01:25
Breath Volume	4.250	01:25
Air Blank	0.000	01:25
Control Test	0.079	01:26
Air Blank	0.000	01:26
Diagnostics Check	OK	01:26

Data Download Date/Time: 02/25/2016 12:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,MICHAEL,L
BREATH TEST OPERATOR AGENCY: Fort Pierce PD
ARRESTING OFFICER: FRAZIER,MICHAEL
ARRESTING OFFICER AGENCY: Fort Pierce PD
LAST AGENCY INSPECTION: 01/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26614080A7
DRY GAS EXPIRY DATE: 10/05/2016
VIOLATION CODE: DUI
ARREST TIME: 21:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:15

SUBJECT'S LAST NAME: NEFF

DATE/TIME OF TEST: 01/21/2016 23:37:35

Diagnostics Check	OK	23:29
Air Blank	0.000	23:30
Control Test	0.080	23:30
Air Blank	0.000	23:30
Subject Sample # 1	0.197	23:32
Breath Volume	1.457	23:32
Air Blank	0.000	23:33
Air Blank	0.000	23:34
Subject Sample # 2	0.203	23:36
Breath Volume	2.601	23:36
Air Blank	0.000	23:36
Control Test	0.079	23:36
Air Blank	0.000	23:37
Diagnostics Check	OK	23:37

Data Download Date/Time: 02/25/2016 12:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEELE,MARK,T
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: STEELE,MARK
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 01/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26614080A7
DRY GAS EXPIRY DATE: 10/05/2016
VIOLATION CODE: DUI
ARREST TIME: 03:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: SAVIORY

DATE/TIME OF TEST: 01/22/2016 03:55:49

Diagnostics Check	OK	03:49
Air Blank	0.000	03:50
Control Test	0.080	03:50
Air Blank	0.000	03:51
Subject Sample # 1	0.172	03:51
Breath Volume	2.058	03:51
Air Blank	0.000	03:52
Air Blank	0.000	03:53
Subject Sample # 2	0.172	03:54
Breath Volume	2.203	03:54
Air Blank	0.000	03:54
Control Test	0.080	03:55
Air Blank	0.000	03:55
Diagnostics Check	OK	03:55

Data Download Date/Time: 02/25/2016 12:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STOKES,SHAYNE,R
BREATH TEST OPERATOR AGENCY: Fort Pierce PD
ARRESTING OFFICER: TIMMONS,BRIAN
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 01/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26614080A7
DRY GAS EXPIRY DATE: 10/05/2016
VIOLATION CODE: DUI
ARREST TIME: 15:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:05

SUBJECT'S LAST NAME: MACK JR

DATE/TIME OF TEST: 01/22/2016 15:56:18

Diagnostics Check	OK	15:50
Air Blank	0.000	15:50
Control Test	0.079	15:50
Air Blank	0.000	15:51
Subject Sample # 1	0.182	15:51
Breath Volume	2.929	15:51
Air Blank	0.000	15:52
Air Blank	0.000	15:54
Subject Sample # 2	0.194	15:54
Breath Volume	3.257	15:54
Air Blank	0.000	15:55
Control Test	0.079	15:55
Air Blank	0.000	15:56
Diagnostics Check	OK	15:56

Data Download Date/Time: 02/25/2016 12:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALVAREZ,JUAN,R
BREATH TEST OPERATOR AGENCY: Fort Pierce PD
ARRESTING OFFICER: FRAZIER,MICHAEL
ARRESTING OFFICER AGENCY: Fort Pierce PD
LAST AGENCY INSPECTION: 01/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26614080A7
DRY GAS EXPIRY DATE: 10/05/2016
VIOLATION CODE: DUI
ARREST TIME: 20:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:07

SUBJECT'S LAST NAME: RAMIREZ

DATE/TIME OF TEST: 01/22/2016 21:33:59

Diagnostics Check	OK	21:33	
Air Blank		21:33	RFI Detect
Air Blank	0.000	21:33	

Data Download Date/Time: 02/25/2016 12:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALVAREZ,JUAN,R
BREATH TEST OPERATOR AGENCY: Fort Pierce PD
ARRESTING OFFICER: FRAIZER,MICHAEL
ARRESTING OFFICER AGENCY: Fort Pierce PD
LAST AGENCY INSPECTION: 01/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26614080A7
DRY GAS EXPIRY DATE: 10/05/2016
VIOLATION CODE: DUI
ARREST TIME: 20:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:07

SUBJECT'S LAST NAME: RAMIREZ

DATE/TIME OF TEST: 01/22/2016 21:43:09

Diagnostics Check	OK	21:36
Air Blank	0.000	21:36
Control Test	0.080	21:37
Air Blank	0.000	21:37
Subject Sample # 1	0.204	21:38
Breath Volume	2.492	21:38
Air Blank	0.000	21:39
Air Blank	0.000	21:41
Subject Sample # 2	0.203	21:41
Breath Volume	1.191	21:41
Air Blank	0.000	21:42
Control Test	0.080	21:42
Air Blank	0.000	21:43
Diagnostics Check	OK	21:43

Data Download Date/Time: 02/25/2016 12:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOGAN,TODD,A
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: HOGAN,TODD
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 01/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26614080A7
DRY GAS EXPIRY DATE: 10/05/2016
VIOLATION CODE: DUI
ARREST TIME: 02:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:47

SUBJECT'S LAST NAME: MOORE

DATE/TIME OF TEST: 01/23/2016 03:15:30

Diagnostics Check	OK	03:08
Air Blank	0.000	03:09
Control Test	0.080	03:09
Air Blank	0.000	03:10
Subject Sample # 1	0.058	03:10
Breath Volume	2.914	03:10
Air Blank	0.000	03:11
Air Blank	0.000	03:13
Subject Sample # 2	0.055	03:14
Breath Volume	3.191	03:14
Air Blank	0.000	03:14
Control Test	0.080	03:14
Air Blank	0.000	03:15
Diagnostics Check	OK	03:15

Data Download Date/Time: 02/25/2016 12:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FELIX,GEORGE,D
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: FELIX,GEORGE
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 01/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26614080A7
DRY GAS EXPIRY DATE: 10/05/2016
VIOLATION CODE: DUI
ARREST TIME: 22:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:21

SUBJECT'S LAST NAME: LOPEZ MORENO

DATE/TIME OF TEST: 01/23/2016 23:49:20

Diagnostics Check	OK	23:42
Air Blank	0.000	23:42
Control Test	0.080	23:42
Air Blank	0.000	23:43
Subject Sample # 1	0.157	23:44
Breath Volume	1.199	23:44
Air Blank	0.000	23:45
Air Blank	0.000	23:47
Subject Sample # 2	0.161	23:47
Breath Volume	1.722	23:47
Air Blank	0.000	23:48
Control Test	0.080	23:48
Air Blank	0.000	23:49
Diagnostics Check	OK	23:49

Data Download Date/Time: 02/25/2016 12:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DRAWDY,LAINE,A
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: DRAWDY,LAINE
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 01/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26614080A7
DRY GAS EXPIRY DATE: 10/05/2016
VIOLATION CODE: DUI
ARREST TIME: 23:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:11

SUBJECT'S LAST NAME: ALESSANDRO

DATE/TIME OF TEST: 01/24/2016 00:45:34

Diagnostics Check	OK	00:38
Air Blank	0.000	00:39
Control Test	0.080	00:39
Air Blank	0.000	00:39
Subject Sample # 1	0.138	00:40
Breath Volume	2.144	00:40
Air Blank	0.000	00:41
Air Blank	0.000	00:43
Subject Sample # 2	0.120	00:44
Breath Volume	1.738	00:44
Air Blank	0.000	00:44
Control Test	0.080	00:44
Air Blank	0.000	00:45
Diagnostics Check	OK	00:45

Data Download Date/Time: 02/25/2016 12:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOGAN,TODD,A
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: HOGAN,TODD
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 01/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26614080A7
DRY GAS EXPIRY DATE: 10/05/2016
VIOLATION CODE: DUI
ARREST TIME: 01:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: BETURNEY

DATE/TIME OF TEST: 01/24/2016 02:23:10

Diagnostics Check	OK	02:17
Air Blank	0.000	02:17
Control Test	0.081	02:17
Air Blank	0.000	02:18
Subject Sample # 1	0.091	02:18
Breath Volume	1.996	02:18
Air Blank	0.000	02:19
Air Blank	0.000	02:21
Subject Sample # 2	0.088	02:21
Breath Volume	1.507	02:21
Air Blank	0.000	02:22
Control Test	0.080	02:22
Air Blank	0.000	02:23
Diagnostics Check	OK	02:23

Data Download Date/Time: 02/25/2016 12:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,MICHAEL,L
BREATH TEST OPERATOR AGENCY: Fort Pierce PD
ARRESTING OFFICER: WACHENDORFER,JEFFREY
ARRESTING OFFICER AGENCY: Fort Pierce PD
LAST AGENCY INSPECTION: 01/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26614080A7
DRY GAS EXPIRY DATE: 10/05/2016
VIOLATION CODE: DUI
ARREST TIME: 01:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:27

SUBJECT'S LAST NAME: TINOCO

DATE/TIME OF TEST: 01/24/2016 02:54:07

Diagnostics Check	OK	02:47
Air Blank	0.000	02:48
Control Test	0.081	02:48
Air Blank	0.000	02:49
Subject Sample # 1	0.237	02:49
Breath Volume	2.648	02:49
Air Blank	0.000	02:50
Air Blank	0.000	02:52
Subject Sample # 2	0.227	02:52
Breath Volume	1.101	02:52
Air Blank	0.000	02:53
Control Test	0.080	02:53
Air Blank	0.000	02:54
Diagnostics Check	OK	02:54

Data Download Date/Time: 02/25/2016 12:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAINTILIE,BOB,
BREATH TEST OPERATOR AGENCY: FHP Troop K
ARRESTING OFFICER: SAINTILIE,BOB
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 01/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26614080A7
DRY GAS EXPIRY DATE: 10/05/2016
VIOLATION CODE: DUI
ARREST TIME: 21:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:17

SUBJECT'S LAST NAME: MARTINO CASTRO

DATE/TIME OF TEST: 01/25/2016 00:07:38

Diagnostics Check	OK	00:04	
Air Blank	0.000	00:05	
Control Test	0.081	00:05	
Air Blank	0.000	00:06	
Subject Sample	# 1	00:06	Subject Test Refused
Breath Volume	0.000	00:06	
Air Blank	0.000	00:06	
Control Test	0.081	00:07	
Air Blank	0.000	00:07	
Diagnostics Check	OK	00:07	

Data Download Date/Time: 02/25/2016 12:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FELIX,GEORGE,D
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: FELIX,GEORGE
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 01/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26614080A7
DRY GAS EXPIRY DATE: 10/05/2016
VIOLATION CODE: DUI
ARREST TIME: 03:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: CHIANESE

DATE/TIME OF TEST: 01/25/2016 04:46:44

Diagnostics Check	OK	04:40
Air Blank	0.000	04:40
Control Test	0.081	04:40
Air Blank	0.000	04:41
Subject Sample # 1	0.114	04:42
Breath Volume	2.460	04:42
Air Blank	0.000	04:42
Air Blank	0.000	04:44
Subject Sample # 2	0.107	04:45
Breath Volume	2.460	04:45
Air Blank	0.000	04:45
Control Test	0.081	04:46
Air Blank	0.000	04:46
Diagnostics Check	OK	04:46

Data Download Date/Time: 02/25/2016 12:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAINTILIE,BOB,
BREATH TEST OPERATOR AGENCY: FHP Troop K
ARRESTING OFFICER: SAINTILIE,BOB
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 01/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26614080A7
DRY GAS EXPIRY DATE: 10/05/2016
VIOLATION CODE: DUI
ARREST TIME: 21:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:19

SUBJECT'S LAST NAME: SCHMIDT

DATE/TIME OF TEST: 01/28/2016 21:50:14

Diagnostics Check	OK	21:40	
Air Blank	0.000	21:40	
Control Test	0.080	21:40	
Air Blank	0.000	21:41	
Subject Sample	# 1 0.178	21:43	
Breath Volume	2.769	21:43	
Air Blank	0.000	21:44	
Air Blank	0.000	21:46	
Subject Sample	# 2	21:48	RFI Detect
Breath Volume	1.109	21:48	
Air Blank	0.000	21:49	
Control Test	0.074	21:49	Control Outside Tolerance
Air Blank	0.000	21:50	

Data Download Date/Time: 02/25/2016 12:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOGAN,TODD,A
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: HOGAN,TODD
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 01/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26614080A7
DRY GAS EXPIRY DATE: 10/05/2016
VIOLATION CODE: DUI
ARREST TIME: 02:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:07

SUBJECT'S LAST NAME: PERKINS

DATE/TIME OF TEST: 01/29/2016 03:39:44

Diagnostics Check	OK	03:33
Air Blank	0.000	03:34
Control Test	0.079	03:34
Air Blank	0.000	03:34
Subject Sample # 1	0.250	03:35
Breath Volume	2.453	03:35
Air Blank	0.000	03:35
Air Blank	0.000	03:37
Subject Sample # 2	0.253	03:38
Breath Volume	2.570	03:38
Air Blank	0.000	03:38
Control Test	0.079	03:39
Air Blank	0.000	03:39
Diagnostics Check	OK	03:39

Data Download Date/Time: 02/25/2016 12:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FELIX,GEORGE,D
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: FELIX,GEORGE
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 01/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26614080A7
DRY GAS EXPIRY DATE: 10/05/2016
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:49

SUBJECT'S LAST NAME: BURGOS

DATE/TIME OF TEST: 01/31/2016 03:22:10

Diagnostics Check	OK	03:15
Air Blank	0.000	03:15
Control Test	0.080	03:16
Air Blank	0.000	03:16
Subject Sample # 1	0.174	03:17
Breath Volume	2.605	03:17
Air Blank	0.000	03:18
Air Blank	0.000	03:20
Subject Sample # 2	0.173	03:20
Breath Volume	1.636	03:20
Air Blank	0.000	03:21
Control Test	0.080	03:21
Air Blank	0.000	03:22
Diagnostics Check	OK	03:22

Data Download Date/Time: 02/25/2016 12:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000795

INSTRUMENT REGISTERED TO: GULF COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURCH,RUSSELL,L
BREATH TEST OPERATOR AGENCY: Port St. Joe PD
ARRESTING OFFICER: BURCH,RUSSELL
ARRESTING OFFICER AGENCY: Port St. Joe PD
LAST AGENCY INSPECTION: 12/21/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 18:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:24

SUBJECT'S LAST NAME: MARTIN

DATE/TIME OF TEST: 01/22/2016 18:56:28

Diagnostics Check	OK	18:50
Air Blank	0.000	18:50
Control Test	0.078	18:51
Air Blank	0.000	18:51
Subject Sample # 1	0.321	18:52
Breath Volume	1.121	18:52
Air Blank	0.000	18:52
Air Blank	0.000	18:54
Subject Sample # 2	0.310	18:55
Breath Volume	1.300	18:55
Air Blank	0.000	18:55
Control Test	0.078	18:55
Air Blank	0.000	18:56
Diagnostics Check	OK	18:56

Data Download Date/Time: 01/29/2016 11:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MALONE, C
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/10/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:04

SUBJECT'S LAST NAME: BANEGAS

DATE/TIME OF TEST: 01/01/2016 03:38:31

Diagnostics Check	OK	03:24	
Air Blank	0.000	03:25	
Control Test	0.080	03:25	
Air Blank	0.000	03:25	
Subject Sample	# 1 0.112	03:29	Volume Not Met
Breath Volume	0.085	03:29	
Air Blank	0.000	03:29	
Air Blank	0.000	03:31	
Subject Sample	# 2 0.178	03:33	
Breath Volume	2.417	03:33	
Air Blank	0.000	03:34	
Air Blank	0.000	03:36	
Subject Sample	# 3 0.162	03:36	
Breath Volume	1.460	03:36	
Air Blank	0.000	03:37	
Control Test	0.077	03:37	
Air Blank	0.000	03:38	
Diagnostics Check	OK	03:38	

Data Download Date/Time: 01/04/2016 19:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SARRO,TAYLOR,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 12/10/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:02

SUBJECT'S LAST NAME: CROY

DATE/TIME OF TEST: 01/01/2016 05:28:41

Diagnostics Check	OK	05:22
Air Blank	0.000	05:22
Control Test	0.080	05:23
Air Blank	0.000	05:23
Subject Sample # 1	0.109	05:23
Breath Volume	1.167	05:23
Air Blank	0.000	05:24
Air Blank	0.000	05:26
Subject Sample # 2	0.120	05:27
Breath Volume	3.011	05:27
Air Blank	0.000	05:27
Control Test	0.080	05:28
Air Blank	0.000	05:28
Diagnostics Check	OK	05:28

Data Download Date/Time: 01/04/2016 19:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY,PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WOODEN,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/10/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 19:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:37

SUBJECT'S LAST NAME: FRANCOIS

DATE/TIME OF TEST: 01/01/2016 20:12:05

Diagnostics Check	OK	19:59	
Air Blank	0.000	19:59	
Control Test	0.079	20:00	
Air Blank	0.000	20:00	
Subject Sample	# 1 0.199	20:03	Volume Not Met
Breath Volume	0.125	20:03	
Air Blank	0.000	20:04	
Air Blank	0.000	20:06	
Subject Sample	# 2 0.214	20:07	
Breath Volume	1.843	20:07	
Air Blank	0.000	20:08	
Air Blank	0.000	20:10	
Subject Sample	# 3 0.223	20:10	
Breath Volume	1.433	20:10	
Air Blank	0.000	20:11	
Control Test	0.078	20:11	
Air Blank	0.000	20:12	
Diagnostics Check	OK	20:12	

Data Download Date/Time: 01/04/2016 19:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PORTMAN, A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 12/10/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:03

SUBJECT'S LAST NAME: CORREDERA

DATE/TIME OF TEST: 01/02/2016 02:29:52

Diagnostics Check	OK	02:23
Air Blank	0.000	02:24
Control Test	0.079	02:24
Air Blank	0.000	02:24
Subject Sample # 1	0.167	02:25
Breath Volume	1.816	02:25
Air Blank	0.000	02:25
Air Blank	0.000	02:27
Subject Sample # 2	0.169	02:28
Breath Volume	1.296	02:28
Air Blank	0.000	02:28
Control Test	0.079	02:29
Air Blank	0.000	02:29
Diagnostics Check	OK	02:29

Data Download Date/Time: 01/04/2016 19:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HERNANDEZ, A
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/10/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: RESENDIZ-BUENROSTRO

DATE/TIME OF TEST: 01/02/2016 04:02:55

Diagnostics Check	OK	03:55
Air Blank	0.000	03:56
Control Test	0.079	03:56
Air Blank	0.000	03:57
Subject Sample # 1	0.241	03:58
Breath Volume	1.980	03:58
Air Blank	0.000	03:58
Air Blank	0.000	04:00
Subject Sample # 2	0.239	04:01
Breath Volume	2.203	04:01
Air Blank	0.000	04:02
Control Test	0.078	04:02
Air Blank	0.000	04:02
Diagnostics Check	OK	04:02

Data Download Date/Time: 01/04/2016 19:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SIMPKINS,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/10/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:59

SUBJECT'S LAST NAME: FERNANDEZ-CRUZ

DATE/TIME OF TEST: 01/03/2016 00:29:22

Diagnostics Check	OK	00:19
Air Blank	0.000	00:19
Control Test	0.080	00:20
Air Blank	0.000	00:20
Subject Sample # 1	0.213	00:23
Breath Volume	1.964	00:23
Air Blank	0.000	00:24
Air Blank	0.000	00:25
Subject Sample # 2	0.196	00:27
Breath Volume	1.835	00:27
Air Blank	0.000	00:28
Control Test	0.078	00:28
Air Blank	0.000	00:29
Diagnostics Check	OK	00:29

Data Download Date/Time: 01/04/2016 19:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BLANCHARD,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 12/10/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:51

SUBJECT'S LAST NAME: MANSOUR

DATE/TIME OF TEST: 01/03/2016 03:17:48

Diagnostics Check	OK	03:11
Air Blank	0.000	03:11
Control Test	0.080	03:12
Air Blank	0.000	03:12
Subject Sample # 1	0.184	03:13
Breath Volume	1.976	03:13
Air Blank	0.000	03:14
Air Blank	0.000	03:15
Subject Sample # 2	0.176	03:16
Breath Volume	2.160	03:16
Air Blank	0.000	03:16
Control Test	0.079	03:17
Air Blank	0.000	03:17
Diagnostics Check	OK	03:17

Data Download Date/Time: 01/04/2016 19:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: CRUZ,E
ARRESTING OFFICER AGENCY: Plant City PD
LAST AGENCY INSPECTION: 12/10/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:31

SUBJECT'S LAST NAME: WASHINGTON

DATE/TIME OF TEST: 01/03/2016 05:57:49

Diagnostics Check	OK	05:51
Air Blank	0.000	05:51
Control Test	0.079	05:52
Air Blank	0.000	05:52
Subject Sample # 1	0.034	05:53
Breath Volume	1.542	05:53
Air Blank	0.000	05:53
Air Blank	0.000	05:55
Subject Sample # 2	0.034	05:56
Breath Volume	1.355	05:56
Air Blank	0.000	05:56
Control Test	0.079	05:57
Air Blank	0.000	05:57
Diagnostics Check	OK	05:57

Data Download Date/Time: 01/04/2016 19:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MALONE,C
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/10/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:18

SUBJECT'S LAST NAME: FAIRCLOTH

DATE/TIME OF TEST: 01/03/2016 23:45:13

Diagnostics Check	OK	23:38
Air Blank	0.000	23:38
Control Test	0.080	23:39
Air Blank	0.000	23:39
Subject Sample # 1	0.000	23:40
Breath Volume	2.164	23:40
Air Blank	0.000	23:41
Air Blank	0.000	23:43
Subject Sample # 2	0.000	23:43
Breath Volume	2.203	23:43
Air Blank	0.000	23:44
Control Test	0.080	23:44
Air Blank	0.000	23:45
Diagnostics Check	OK	23:45

Data Download Date/Time: 01/04/2016 19:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VAN-TREESE,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 12/10/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:33

SUBJECT'S LAST NAME: COLLAZO VAZQUEZ

DATE/TIME OF TEST: 01/04/2016 00:59:56

Diagnostics Check	OK	00:53
Air Blank	0.000	00:54
Control Test	0.079	00:54
Air Blank	0.000	00:54
Subject Sample # 1	0.153	00:55
Breath Volume	2.656	00:55
Air Blank	0.000	00:56
Air Blank	0.000	00:57
Subject Sample # 2	0.145	00:58
Breath Volume	2.917	00:58
Air Blank	0.000	00:59
Control Test	0.079	00:59
Air Blank	0.000	00:59
Diagnostics Check	OK	00:59

Data Download Date/Time: 01/04/2016 19:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASTELLANO,BRENT,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: CASTELLANO,BRENT
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 01/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1766638
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: BENSON

DATE/TIME OF TEST: 01/22/2016 01:52:15

Diagnostics Check	OK	01:49	
Air Blank	0.000	01:49	
Control Test	0.080	01:50	
Air Blank	0.000	01:50	
Subject Sample	# 1	01:50	Subject Test Refused
Breath Volume	0.000	01:50	
Air Blank	0.000	01:51	
Control Test	0.080	01:51	
Air Blank	0.000	01:52	
Diagnostics Check	OK	01:52	

Data Download Date/Time: 02/17/2016 15:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,GARY,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: MILLER,GARY
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 01/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1766638
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:09

SUBJECT'S LAST NAME: REINAGEL

DATE/TIME OF TEST: 01/23/2016 02:35:38

Diagnostics Check	OK	02:29
Air Blank	0.000	02:29
Control Test	0.080	02:30
Air Blank	0.000	02:30
Subject Sample # 1	0.128	02:31
Breath Volume	1.324	02:31
Air Blank	0.000	02:31
Air Blank	0.000	02:33
Subject Sample # 2	0.127	02:34
Breath Volume	1.359	02:34
Air Blank	0.000	02:34
Control Test	0.081	02:35
Air Blank	0.000	02:35
Diagnostics Check	OK	02:35

Data Download Date/Time: 02/17/2016 15:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SERRANO,E,
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: MASLYN,J
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 01/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1662060
DRY GAS EXPIRY DATE: 04/18/2017
VIOLATION CODE: DUI
ARREST TIME: 11:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:18

SUBJECT'S LAST NAME: GOSNELL

DATE/TIME OF TEST: 01/24/2016 00:54:41

Diagnostics Check	OK	00:44
Air Blank	0.000	00:45
Control Test	0.080	00:45
Air Blank	0.000	00:46
Subject Sample # 1	0.240	00:47
Breath Volume	1.558	00:47
Air Blank	0.000	00:48
Air Blank	0.000	00:49
Subject Sample # 2	0.214	00:50
Breath Volume	1.894	00:50
Air Blank	0.000	00:50
Air Blank	0.000	00:52
Subject Sample # 3	0.236	00:53
Breath Volume	1.160	00:53
Air Blank	0.000	00:53
Control Test	0.082	00:54
Air Blank	0.000	00:54
Diagnostics Check	OK	00:54

Data Download Date/Time: 02/17/2016 15:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,MANUEL,A
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: HERNANDEZ,MANUEL
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 01/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1766638
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:42

SUBJECT'S LAST NAME: FEAGIN

DATE/TIME OF TEST: 01/24/2016 04:06:31

Diagnostics Check	OK	04:03	
Air Blank	0.000	04:03	
Control Test	0.080	04:04	
Air Blank	0.000	04:04	
Subject Sample	# 1	04:05	Subject Test Refused
Breath Volume	0.000	04:05	
Air Blank	0.000	04:05	
Control Test	0.081	04:05	
Air Blank	0.000	04:06	
Diagnostics Check	OK	04:06	

Data Download Date/Time: 02/17/2016 15:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GAGO-
CHUQUILLANQUI,ALBERTO,E
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: GAGO-
CHUQUILLANQUI,ALBERTO
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 01/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1766638
DRY GAS EXPIRY DATE: 11/05/2017

VIOLATION CODE: DUI
ARREST TIME: 19:56
GENDER: Male

OBSERVATION PERIOD BEGAN: 20:04

SUBJECT'S LAST NAME: HOLIDAY

DATE/TIME OF TEST: 01/24/2016 21:36:51

Diagnostics Check	OK	21:29
Air Blank	0.000	21:29
Control Test	0.081	21:30
Air Blank	0.000	21:30
Subject Sample # 1	0.192	21:31
Breath Volume	2.363	21:31
Air Blank	0.000	21:32
Air Blank	0.000	21:34
Subject Sample # 2	0.200	21:35
Breath Volume	2.492	21:35
Air Blank	0.000	21:35
Control Test	0.082	21:36
Air Blank	0.000	21:36
Diagnostics Check	OK	21:36

Data Download Date/Time: 02/17/2016 15:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOBOSIEWICZ, ANTHONY, N
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: WOOD, M
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 01/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1766638
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 15:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:39

SUBJECT'S LAST NAME: MOSS

DATE/TIME OF TEST: 01/25/2016 17:24:59

Diagnostics Check	OK	17:18
Air Blank	0.000	17:18
Control Test	0.081	17:18
Air Blank	0.000	17:19
Subject Sample # 1	0.139	17:19
Breath Volume	1.292	17:19
Air Blank	0.000	17:20
Air Blank	0.000	17:22
Subject Sample # 2	0.138	17:23
Breath Volume	1.964	17:23
Air Blank	0.000	17:24
Control Test	0.083	17:24
Air Blank	0.000	17:24
Diagnostics Check	OK	17:24

Data Download Date/Time: 02/17/2016 15:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOOREHEAD,DAVID,L
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: MOOREHEAD,DAVID
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 01/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1766638
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:32

SUBJECT'S LAST NAME: JOSE CABRERA

DATE/TIME OF TEST: 01/27/2016 01:09:50

Diagnostics Check	OK	00:57	
Air Blank	0.000	00:58	
Control Test	0.079	00:58	
Air Blank	0.000	00:59	
Subject Sample	# 1 0.175	01:02	Volume Not Met
Breath Volume	0.238	01:02	
Air Blank	0.000	01:03	
Air Blank	0.000	01:04	
Subject Sample	# 2 0.176	01:08	Volume Not Met
Breath Volume	0.031	01:08	
Air Blank	0.000	01:08	
Control Test	0.080	01:09	
Air Blank	0.000	01:09	
Diagnostics Check	OK	01:09	

Data Download Date/Time: 02/17/2016 15:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOOREHEAD,DAVID,L
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: MOOREHEAD,DAVID
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 01/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1766638
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:54

SUBJECT'S LAST NAME: JOSE CABRERA

DATE/TIME OF TEST: 01/27/2016 03:20:45

Diagnostics Check	OK	03:14	
Air Blank	0.000	03:15	
Control Test	0.079	03:15	
Air Blank	0.000	03:16	
Subject Sample	# 1 0.109	03:19	Volume Not Met
Breath Volume	0.421	03:19	
Air Blank		03:19	Purge Fail
Air Blank	0.000	03:20	

Data Download Date/Time: 02/17/2016 15:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOOREHEAD,DAVID,L
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: MOOREHEAD,DAVID
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 01/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1766638
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:54

SUBJECT'S LAST NAME: JOSE CABRERA

DATE/TIME OF TEST: 01/27/2016 03:31:03

Diagnostics Check	OK	03:22
Air Blank	0.000	03:23
Control Test	0.079	03:23
Air Blank	0.000	03:23
Subject Sample # 1	0.122	03:24
Breath Volume	1.496	03:24
Air Blank	0.000	03:24
Air Blank	0.000	03:26
Subject Sample # 2	0.119	03:29
Breath Volume	1.457	03:29
Air Blank	0.000	03:30
Control Test	0.082	03:30
Air Blank	0.000	03:30
Diagnostics Check	OK	03:31

Data Download Date/Time: 02/17/2016 15:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERNANDEZ,ALAIN,
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: FERNANDEZ,ALAIN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 01/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1766638
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: BUNDSCHUH

DATE/TIME OF TEST: 01/28/2016 03:58:07

Diagnostics Check	OK	03:55	
Air Blank	0.000	03:56	
Control Test	0.080	03:56	
Air Blank	0.000	03:56	
Subject Sample	# 1	03:57	Subject Test Refused
Breath Volume	0.000	03:57	
Air Blank		03:57	RFI Detect
Air Blank	0.000	03:58	

Data Download Date/Time: 02/17/2016 15:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERNANDEZ,ALAIN,
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: FERNANDEZ,ALAIN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 01/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1766638
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: B532753663760

DATE/TIME OF TEST: 01/28/2016 04:03:46

Diagnostics Check	OK	04:00	
Air Blank	0.000	04:01	
Control Test	0.080	04:01	
Air Blank	0.000	04:02	
Subject Sample	# 1	04:02	Subject Test Refused
Breath Volume	0.000	04:02	
Air Blank	0.000	04:02	
Control Test	0.080	04:03	
Air Blank	0.000	04:03	
Diagnostics Check	OK	04:03	

Data Download Date/Time: 02/17/2016 15:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERNANDEZ,ALAIN,
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: FERNANDEZ,ALAIN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 01/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1766638
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:37

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 01/29/2016 01:00:39

Diagnostics Check	OK	00:57	
Air Blank	0.000	00:58	
Control Test	0.080	00:58	
Air Blank	0.000	00:59	
Subject Sample	# 1	00:59	Subject Test Refused
Breath Volume	0.000	00:59	
Air Blank	0.000	00:59	
Control Test	0.080	01:00	
Air Blank	0.000	01:00	
Diagnostics Check	OK	01:00	

Data Download Date/Time: 02/17/2016 15:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,GARY,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: MILLER,GARY
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 01/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1766638
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: RILEY

DATE/TIME OF TEST: 01/30/2016 23:34:24

Diagnostics Check	OK	23:31	
Air Blank	0.000	23:32	
Control Test	0.080	23:32	
Air Blank	0.000	23:32	
Subject Sample	# 1	23:33	RFI Detect
Breath Volume	0.000	23:33	
Air Blank	0.000	23:33	
Control Test	0.080	23:33	
Air Blank	0.000	23:34	
Diagnostics Check	OK	23:34	

Data Download Date/Time: 02/17/2016 15:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,GARY,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: MILLER,GARY
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 01/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1766638
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: RILEY

DATE/TIME OF TEST: 01/30/2016 23:39:23

Diagnostics Check	OK	23:35	
Air Blank	0.000	23:36	
Control Test	0.081	23:36	
Air Blank	0.000	23:37	
Subject Sample	# 1	23:38	Interferent Detect
Breath Volume	1.468	23:38	
Air Blank		23:38	RFI Detect
Air Blank		23:39	Purge Fail

Data Download Date/Time: 02/17/2016 15:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUSTAMANTE,EDUARDO,B
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: BUSTAMANTE,EDUARDO
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 01/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1662060
DRY GAS EXPIRY DATE: 04/18/2017
VIOLATION CODE: DUI
ARREST TIME: 22:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:29

SUBJECT'S LAST NAME: KOTZIAN

DATE/TIME OF TEST: 01/31/2016 00:10:31

Diagnostics Check	OK	00:04
Air Blank	0.000	00:04
Control Test	0.079	00:05
Air Blank	0.000	00:05
Subject Sample # 1	0.221	00:05
Breath Volume	3.199	00:05
Air Blank	0.000	00:06
Air Blank	0.000	00:08
Subject Sample # 2	0.220	00:08
Breath Volume	4.300	00:08
Air Blank	0.000	00:09
Control Test	0.080	00:09
Air Blank	0.000	00:10
Diagnostics Check	OK	00:10

Data Download Date/Time: 02/17/2016 15:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WIGGLESWORTH,JOE,B
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: BUTLER,E
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 01/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28814080A3
DRY GAS EXPIRY DATE: 11/05/2016
VIOLATION CODE: DUI
ARREST TIME: 20:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:29

SUBJECT'S LAST NAME: COWART

DATE/TIME OF TEST: 01/30/2016 20:56:44

Diagnostics Check	OK	20:49
Air Blank	0.000	20:50
Control Test	0.078	20:50
Air Blank	0.000	20:51
Subject Sample # 1	0.195	20:52
Breath Volume	2.187	20:52
Air Blank	0.000	20:52
Air Blank	0.000	20:54
Subject Sample # 2	0.194	20:55
Breath Volume	2.132	20:55
Air Blank	0.000	20:55
Control Test	0.078	20:56
Air Blank	0.000	20:56
Diagnostics Check	OK	20:56

Data Download Date/Time: 02/13/2016 19:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FORINASH,GARY,B
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: RAFFERTY,C
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 01/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28814080A3
DRY GAS EXPIRY DATE: 11/05/2016
VIOLATION CODE: DUI
ARREST TIME: 19:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:04

SUBJECT'S LAST NAME: TESTA

DATE/TIME OF TEST: 01/30/2016 21:32:42

Diagnostics Check	OK	21:24
Air Blank	0.000	21:25
Control Test	0.078	21:25
Air Blank	0.000	21:26
Subject Sample # 1	0.181	21:27
Breath Volume	1.328	21:27
Air Blank	0.000	21:28
Air Blank	0.000	21:29
Subject Sample # 2	0.174	21:30
Breath Volume	1.921	21:30
Air Blank	0.000	21:31
Control Test	0.077	21:32
Air Blank	0.000	21:32
Diagnostics Check	OK	21:32

Data Download Date/Time: 02/13/2016 19:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WIGGLESWORTH,JOE,B
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: TAN,R
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 01/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28814080A3
DRY GAS EXPIRY DATE: 11/05/2016
VIOLATION CODE: DUI
ARREST TIME: 23:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:42

SUBJECT'S LAST NAME: SHEARON

DATE/TIME OF TEST: 01/31/2016 01:07:48

Diagnostics Check	OK	01:04	
Air Blank	0.000	01:05	
Control Test	0.079	01:05	
Air Blank	0.000	01:06	
Subject Sample	# 1	01:06	Subject Test Refused
Breath Volume	0.000	01:06	
Air Blank	0.000	01:06	
Control Test	0.078	01:07	
Air Blank	0.000	01:07	
Diagnostics Check	OK	01:07	

Data Download Date/Time: 02/13/2016 19:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARSHALL,JARROD,C
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: ORDER,COURT
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 12/31/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28814080A3
DRY GAS EXPIRY DATE: 11/05/2016
VIOLATION CODE: DUI
ARREST TIME: 12:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 12:38

SUBJECT'S LAST NAME: CARUTHERS

DATE/TIME OF TEST: 01/01/2016 13:06:10

Diagnostics Check	OK	13:00
Air Blank	0.000	13:00
Control Test	0.077	13:00
Air Blank	0.000	13:01
Subject Sample # 1	0.000	13:01
Breath Volume	1.222	13:01
Air Blank	0.000	13:02
Air Blank	0.000	13:04
Subject Sample # 2	0.000	13:04
Breath Volume	1.722	13:04
Air Blank	0.000	13:05
Control Test	0.077	13:05
Air Blank	0.000	13:06
Diagnostics Check	OK	13:06

Data Download Date/Time: 01/30/2016 15:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLIS,TERENCE,L
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: PETERSON,ROBERT
ARRESTING OFFICER AGENCY: Wildwood PD
LAST AGENCY INSPECTION: 12/31/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28814080A3
DRY GAS EXPIRY DATE: 11/05/2016
VIOLATION CODE: DUI
ARREST TIME: 03:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:43

SUBJECT'S LAST NAME: POWELL

DATE/TIME OF TEST: 01/04/2016 04:19:01

Diagnostics Check	OK	04:10
Air Blank	0.000	04:10
Control Test	0.079	04:11
Air Blank	0.000	04:11
Subject Sample # 1	0.000	04:14
Breath Volume	1.128	04:14
Air Blank	0.000	04:14
Air Blank	0.000	04:16
Subject Sample # 2	0.000	04:17
Breath Volume	1.164	04:17
Air Blank	0.000	04:18
Control Test	0.078	04:18
Air Blank	0.000	04:18
Diagnostics Check	OK	04:19

Data Download Date/Time: 01/30/2016 15:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DICKERSON,AMELIA,D
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: SEARS,T
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 12/31/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28814080A3
DRY GAS EXPIRY DATE: 11/05/2016
VIOLATION CODE: DUI
ARREST TIME: 20:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: SULLIVAN

DATE/TIME OF TEST: 01/22/2016 22:27:58

Diagnostics Check	OK	22:20
Air Blank	0.000	22:21
Control Test	0.078	22:21
Air Blank	0.000	22:22
Subject Sample # 1	0.000	22:23
Breath Volume	2.949	22:23
Air Blank	0.000	22:24
Air Blank	0.000	22:26
Subject Sample # 2	0.000	22:26
Breath Volume	1.785	22:26
Air Blank	0.000	22:27
Control Test	0.077	22:27
Air Blank	0.000	22:27
Diagnostics Check	OK	22:27

Data Download Date/Time: 01/30/2016 15:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DICKERSON, AMELIA, D
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: BARD, J
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/31/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28814080A3
DRY GAS EXPIRY DATE: 11/05/2016
VIOLATION CODE: DUI
ARREST TIME: 21:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: KRAPE

DATE/TIME OF TEST: 01/23/2016 23:51:53

Diagnostics Check	OK	23:45
Air Blank	0.000	23:45
Control Test	0.079	23:46
Air Blank	0.000	23:46
Subject Sample # 1	0.181	23:47
Breath Volume	3.156	23:47
Air Blank	0.000	23:47
Air Blank	0.000	23:49
Subject Sample # 2	0.176	23:50
Breath Volume	3.058	23:50
Air Blank	0.000	23:50
Control Test	0.078	23:51
Air Blank	0.000	23:51
Diagnostics Check	OK	23:51

Data Download Date/Time: 01/30/2016 15:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUTTON,EDWIN,C
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: RAFFERTY,C
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 12/31/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28814080A3
DRY GAS EXPIRY DATE: 11/05/2016
VIOLATION CODE: DUI
ARREST TIME: 03:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:14

SUBJECT'S LAST NAME: PACKARD

DATE/TIME OF TEST: 01/27/2016 04:47:13

Diagnostics Check	OK	04:37
Air Blank	0.000	04:37
Control Test	0.077	04:37
Air Blank	0.000	04:38
Subject Sample # 1	0.198	04:39
Breath Volume	1.367	04:39
Air Blank	0.000	04:39
Air Blank	0.000	04:41
Subject Sample # 2	0.228	04:42
Breath Volume	3.136	04:42
Air Blank	0.000	04:43
Air Blank	0.000	04:44
Subject Sample # 3	0.216	04:45
Breath Volume	2.054	04:45
Air Blank	0.000	04:46
Control Test	0.076	04:46
Air Blank	0.000	04:47
Diagnostics Check	OK	04:47

Data Download Date/Time: 01/30/2016 15:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WIGGLESWORTH,JOE,B
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: SARGENT,B
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 12/31/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28814080A3
DRY GAS EXPIRY DATE: 11/05/2016
VIOLATION CODE: DUI
ARREST TIME: 18:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:37

SUBJECT'S LAST NAME: DUKET

DATE/TIME OF TEST: 01/29/2016 20:14:10

Diagnostics Check	OK	20:11	
Air Blank	0.000	20:11	
Control Test	0.079	20:11	
Air Blank	0.000	20:12	
Subject Sample	# 1	20:12	Subject Test Refused
Breath Volume	0.000	20:12	
Air Blank	0.000	20:13	
Control Test	0.078	20:13	
Air Blank	0.000	20:14	
Diagnostics Check	OK	20:14	

Data Download Date/Time: 01/30/2016 15:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WIGGLESWORTH,JOE,B
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: OTERO,H
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 12/31/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28814080A3
DRY GAS EXPIRY DATE: 11/05/2016
VIOLATION CODE: DUI
ARREST TIME: 20:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:58

SUBJECT'S LAST NAME: HEBERT

DATE/TIME OF TEST: 01/29/2016 21:21:41

Diagnostics Check	OK	21:18	
Air Blank	0.000	21:19	
Control Test	0.077	21:19	
Air Blank	0.000	21:20	
Subject Sample	# 1	21:20	Subject Test Refused
Breath Volume	0.000	21:20	
Air Blank	0.000	21:20	
Control Test	0.078	21:21	
Air Blank	0.000	21:21	
Diagnostics Check	OK	21:21	

Data Download Date/Time: 01/30/2016 15:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUTTON, EDWIN, C
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: RAFFERTY, C
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 12/31/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28814080A3
DRY GAS EXPIRY DATE: 11/05/2016
VIOLATION CODE: DUI
ARREST TIME: 03:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:33

SUBJECT'S LAST NAME: BEAUCHEMIN

DATE/TIME OF TEST: 01/30/2016 04:05:57

Diagnostics Check	OK	03:59
Air Blank	0.000	03:59
Control Test	0.077	04:00
Air Blank	0.000	04:00
Subject Sample # 1	0.064	04:01
Breath Volume	3.375	04:01
Air Blank	0.000	04:02
Air Blank	0.000	04:03
Subject Sample # 2	0.059	04:04
Breath Volume	2.148	04:04
Air Blank	0.000	04:05
Control Test	0.078	04:05
Air Blank	0.000	04:05
Diagnostics Check	OK	04:05

Data Download Date/Time: 01/30/2016 15:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHOETTNER,ADRIAN,J
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: VICK,B
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 12/17/2015

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: LEETE

DATE/TIME OF TEST: 01/01/2016 01:37:42

Diagnostics Check	OK	01:34	
Air Blank	0.000	01:35	
Control Test	0.078	01:35	
Air Blank	0.000	01:36	
Subject Sample	# 1	01:36	Subject Test Refused
Breath Volume	0.000	01:36	
Air Blank	0.000	01:36	
Control Test	0.077	01:37	
Air Blank	0.000	01:37	
Diagnostics Check	OK	01:37	

Data Download Date/Time: 01/25/2016 11:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: METZ,KYLE,J
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: GRANGE,ROBERT
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 12/17/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 08:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:02

SUBJECT'S LAST NAME: MARR

DATE/TIME OF TEST: 01/02/2016 10:39:38

Diagnostics Check	OK	10:32
Air Blank	0.000	10:33
Control Test	0.078	10:33
Air Blank	0.000	10:34
Subject Sample # 1	0.000	10:35
Breath Volume	2.164	10:35
Air Blank	0.000	10:35
Air Blank	0.000	10:37
Subject Sample # 2	0.000	10:38
Breath Volume	3.046	10:38
Air Blank	0.000	10:38
Control Test	0.077	10:39
Air Blank	0.000	10:39
Diagnostics Check	OK	10:39

Data Download Date/Time: 01/25/2016 11:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCAZZIO,SEAN,J
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: HOWARD,R
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 12/17/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: WEEKS

DATE/TIME OF TEST: 01/03/2016 22:51:49

Diagnostics Check	OK	22:45
Air Blank	0.000	22:46
Control Test	0.077	22:46
Air Blank	0.000	22:46
Subject Sample # 1	0.206	22:47
Breath Volume	4.816	22:47
Air Blank	0.000	22:48
Air Blank	0.000	22:49
Subject Sample # 2	0.213	22:50
Breath Volume	3.078	22:50
Air Blank	0.000	22:50
Control Test	0.077	22:51
Air Blank	0.000	22:51
Diagnostics Check	OK	22:51

Data Download Date/Time: 01/25/2016 11:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCAZZIO,SEAN,J
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: HOWARD,R
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 12/17/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: MCFADDEN

DATE/TIME OF TEST: 01/04/2016 03:33:34

Diagnostics Check	OK	03:30	
Air Blank	0.000	03:31	
Control Test	0.077	03:31	
Air Blank	0.000	03:32	
Subject Sample	# 1	03:32	Subject Test Refused
Breath Volume	0.000	03:32	
Air Blank	0.000	03:32	
Control Test	0.077	03:32	
Air Blank	0.000	03:33	
Diagnostics Check	OK	03:33	

Data Download Date/Time: 01/25/2016 11:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DADDARIO,ASHLEY,A
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: SUDLOW,J
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 12/17/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 07:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:26

SUBJECT'S LAST NAME: GELLER

DATE/TIME OF TEST: 01/05/2016 13:57:51

Diagnostics Check	OK	13:51
Air Blank	0.000	13:51
Control Test	0.077	13:51
Air Blank	0.000	13:52
Subject Sample # 1	0.007	13:53
Breath Volume	2.109	13:53
Air Blank	0.000	13:53
Air Blank	0.000	13:55
Subject Sample # 2	0.007	13:56
Breath Volume	2.640	13:56
Air Blank	0.000	13:56
Control Test	0.077	13:57
Air Blank	0.000	13:57
Diagnostics Check	OK	13:57

Data Download Date/Time: 01/25/2016 11:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: METZ,KYLE,J
BREATH TEST OPERATOR AGENCY: Citrus County SO
ARRESTING OFFICER: WILLIAMS,DEPUTY
ARRESTING OFFICER AGENCY: CCA Citrus
LAST AGENCY INSPECTION: 12/17/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 15:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:23

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 01/06/2016 15:47:35

Diagnostics Check	OK	15:43	
Air Blank	0.000	15:44	
Control Test	0.077	15:44	
Air Blank	0.000	15:44	
Subject Sample	# 1	15:46	Subject Test Refused
Breath Volume	0.000	15:46	
Air Blank	0.000	15:46	
Control Test	0.078	15:46	
Air Blank	0.000	15:47	
Diagnostics Check	OK	15:47	

Data Download Date/Time: 01/25/2016 11:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHOETTNER,ADRIAN,J
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: AGUILERA,E
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 12/17/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:48

SUBJECT'S LAST NAME: SHERRILL

DATE/TIME OF TEST: 01/07/2016 04:18:02

Diagnostics Check	OK	04:15	
Air Blank	0.000	04:15	
Control Test	0.078	04:15	
Air Blank	0.000	04:16	
Subject Sample	# 1	04:16	Subject Test Refused
Breath Volume	0.000	04:16	
Air Blank	0.000	04:17	
Control Test	0.077	04:17	
Air Blank	0.000	04:17	
Diagnostics Check	OK	04:18	

Data Download Date/Time: 01/25/2016 11:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FOSTER,MICHAEL,D
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: STRICKLAND,D
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 12/17/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 08:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:55

SUBJECT'S LAST NAME: ROBERSON

DATE/TIME OF TEST: 01/10/2016 10:48:15

Diagnostics Check	OK	10:41
Air Blank	0.000	10:41
Control Test	0.077	10:42
Air Blank	0.000	10:42
Subject Sample # 1	0.000	10:43
Breath Volume	2.957	10:43
Air Blank	0.000	10:44
Air Blank	0.000	10:46
Subject Sample # 2	0.000	10:46
Breath Volume	2.328	10:46
Air Blank	0.000	10:47
Control Test	0.078	10:47
Air Blank	0.000	10:48
Diagnostics Check	OK	10:48

Data Download Date/Time: 01/25/2016 11:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCAZZIO,SEAN,J
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: HOWARD,R
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 12/17/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:26

SUBJECT'S LAST NAME: WYATT

DATE/TIME OF TEST: 01/16/2016 03:02:39

Diagnostics Check	OK	02:56
Air Blank	0.000	02:57
Control Test	0.077	02:57
Air Blank	0.000	02:57
Subject Sample # 1	0.202	02:58
Breath Volume	1.996	02:58
Air Blank	0.000	02:58
Air Blank	0.000	03:00
Subject Sample # 2	0.210	03:01
Breath Volume	1.941	03:01
Air Blank	0.000	03:01
Control Test	0.077	03:02
Air Blank	0.000	03:02
Diagnostics Check	OK	03:02

Data Download Date/Time: 01/25/2016 12:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: METZ,KYLE,J
BREATH TEST OPERATOR AGENCY: Citrus County SO
ARRESTING OFFICER: WHITSEL,MICK
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/17/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 14:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:02

SUBJECT'S LAST NAME: PLUMB

DATE/TIME OF TEST: 01/17/2016 16:27:35

Diagnostics Check	OK	16:24	
Air Blank	0.000	16:25	
Control Test	0.077	16:25	
Air Blank	0.000	16:25	
Subject Sample	# 1	16:26	Subject Test Refused
Breath Volume	0.000	16:26	
Air Blank	0.000	16:26	
Control Test	0.077	16:26	
Air Blank	0.000	16:27	
Diagnostics Check	OK	16:27	

Data Download Date/Time: 01/25/2016 12:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOGUE,VINCE,H
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: HORTON,J
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 12/17/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 18:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:50

SUBJECT'S LAST NAME: NOLAN

DATE/TIME OF TEST: 01/19/2016 19:13:55

Diagnostics Check	OK	19:11	
Air Blank	0.000	19:11	
Control Test	0.078	19:11	
Air Blank	0.000	19:12	
Subject Sample	# 1	19:12	Subject Test Refused
Breath Volume	0.000	19:12	
Air Blank	0.000	19:13	
Control Test	0.077	19:13	
Air Blank	0.000	19:13	
Diagnostics Check	OK	19:13	

Data Download Date/Time: 01/25/2016 12:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: METZ,KYLE,J
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: TAYLOR,T
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 01/25/2016

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 13:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:05

SUBJECT'S LAST NAME: RISNER

DATE/TIME OF TEST: 01/25/2016 14:31:59

Diagnostics Check	OK	14:25
Air Blank	0.000	14:26
Control Test	0.077	14:26
Air Blank	0.000	14:27
Subject Sample # 1	0.184	14:27
Breath Volume	2.679	14:27
Air Blank	0.000	14:28
Air Blank	0.000	14:30
Subject Sample # 2	0.179	14:30
Breath Volume	2.976	14:30
Air Blank	0.000	14:31
Control Test	0.077	14:31
Air Blank	0.000	14:31
Diagnostics Check	OK	14:31

Data Download Date/Time: 02/16/2016 14:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHOETTNER,ADRIAN,J
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: REID,W.
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: WHITING

DATE/TIME OF TEST: 01/25/2016 23:53:43

Diagnostics Check	OK	23:50	
Air Blank	0.000	23:51	
Control Test	0.078	23:51	
Air Blank	0.000	23:52	
Subject Sample	# 1	23:52	Subject Test Refused
Breath Volume	0.000	23:52	
Air Blank	0.000	23:52	
Control Test	0.077	23:53	
Air Blank	0.000	23:53	
Diagnostics Check	OK	23:53	

Data Download Date/Time: 02/16/2016 14:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAWRENCE,JEDIDAH,E
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: HUNTER,J
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 01/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 19:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:37

SUBJECT'S LAST NAME: MCKEEHAN

DATE/TIME OF TEST: 01/27/2016 21:22:26

Diagnostics Check	OK	21:15
Air Blank	0.000	21:15
Control Test	0.077	21:16
Air Blank	0.000	21:16
Subject Sample # 1	0.194	21:17
Breath Volume	1.851	21:17
Air Blank	0.000	21:18
Air Blank	0.000	21:20
Subject Sample # 2	0.196	21:20
Breath Volume	1.875	21:20
Air Blank	0.000	21:21
Control Test	0.077	21:21
Air Blank	0.000	21:22
Diagnostics Check	OK	21:22

Data Download Date/Time: 02/16/2016 14:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FOSTER, MICHAEL, D
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: CLOUD, .
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 09:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:16

SUBJECT'S LAST NAME: FALLSTICH

DATE/TIME OF TEST: 01/28/2016 10:44:44

Diagnostics Check	OK	10:36
Air Blank	0.000	10:37
Control Test	0.077	10:37
Air Blank	0.000	10:38
Subject Sample # 1	0.000	10:39
Breath Volume	1.257	10:39
Air Blank	0.000	10:39
Air Blank	0.000	10:41
Subject Sample # 2	0.000	10:43
Breath Volume	1.144	10:43
Air Blank	0.000	10:43
Control Test	0.077	10:44
Air Blank	0.000	10:44
Diagnostics Check	OK	10:44

Data Download Date/Time: 02/16/2016 14:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENTON,AMY,B
BREATH TEST OPERATOR AGENCY: Astatula PD
ARRESTING OFFICER: WHITAKER,NICHAEAL
ARRESTING OFFICER AGENCY: Howey-In-The-Hills
LAST AGENCY INSPECTION: 01/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1766638
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:15

SUBJECT'S LAST NAME: MOORE

DATE/TIME OF TEST: 01/29/2016 22:42:23

Diagnostics Check	OK	22:39	
Air Blank	0.000	22:39	
Control Test	0.079	22:39	
Air Blank	0.000	22:40	
Subject Sample	# 1	22:41	Subject Test Refused
Breath Volume	0.000	22:41	
Air Blank	0.000	22:41	
Control Test	0.079	22:41	
Air Blank	0.000	22:42	
Diagnostics Check	OK	22:42	

Data Download Date/Time: 02/25/2016 10:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENTON,AMY,B
BREATH TEST OPERATOR AGENCY: Astatula PD
ARRESTING OFFICER: GROGAN,KARL
ARRESTING OFFICER AGENCY: Tavares PD
LAST AGENCY INSPECTION: 01/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1766638
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: CLAWSON

DATE/TIME OF TEST: 01/31/2016 23:14:08

Diagnostics Check	OK	23:11	
Air Blank	0.000	23:11	
Control Test	0.079	23:12	
Air Blank	0.000	23:12	
Subject Sample	# 1	23:12	Subject Test Refused
Breath Volume	0.000	23:12	
Air Blank	0.000	23:13	
Control Test	0.079	23:13	
Air Blank	0.000	23:14	
Diagnostics Check	OK	23:14	

Data Download Date/Time: 02/25/2016 10:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ;6360100425072576188,LUIS,M
BREATH TEST OPERATOR AGENCY: FHP Troop K
ARRESTING OFFICER: TEJERA,LUIS
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 12/21/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG402702
DRY GAS EXPIRY DATE: 01/27/2016
VIOLATION CODE: DUI
ARREST TIME: 01:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: DIXON

DATE/TIME OF TEST: 01/02/2016 02:09:00

Diagnostics Check	OK	02:08	
Air Blank		02:08	RFI Detect
Air Blank	0.000	02:09	

Data Download Date/Time: 01/28/2016 22:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORRIVEAU, CHAD, R
BREATH TEST OPERATOR AGENCY: FHP Troop K
ARRESTING OFFICER: OLIVER, CATHERINE
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 12/21/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG402702
DRY GAS EXPIRY DATE: 01/27/2016
VIOLATION CODE: DUI
ARREST TIME: 22:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: ROJAS

DATE/TIME OF TEST: 01/05/2016 23:47:03

Diagnostics Check	OK	23:40
Air Blank	0.000	23:41
Control Test	0.079	23:41
Air Blank	0.000	23:41
Subject Sample # 1	0.114	23:42
Breath Volume	3.242	23:42
Air Blank	0.000	23:43
Air Blank	0.000	23:44
Subject Sample # 2	0.110	23:45
Breath Volume	3.214	23:45
Air Blank	0.000	23:46
Control Test	0.079	23:46
Air Blank	0.000	23:46
Diagnostics Check	OK	23:47

Data Download Date/Time: 01/28/2016 22:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHESSHER,SANDRA,W
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: BLOOM,N
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 12/21/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG402702
DRY GAS EXPIRY DATE: 01/27/2016
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: ABDUL

DATE/TIME OF TEST: 01/09/2016 04:29:15

Diagnostics Check	OK	04:14	
Air Blank	0.000	04:14	
Control Test	0.078	04:15	
Air Blank	0.000	04:15	
Subject Sample	# 1 0.146	04:17	
Breath Volume	1.503	04:17	
Air Blank	0.000	04:18	
Air Blank	0.000	04:20	
Subject Sample	# 2	04:23	No Sample Provided
Breath Volume	0.000	04:23	
Air Blank	0.000	04:23	
Air Blank	0.000	04:25	
Subject Sample	# 3 0.135	04:27	
Breath Volume	1.246	04:27	
Air Blank	0.000	04:28	
Control Test	0.077	04:28	
Air Blank	0.000	04:29	
Diagnostics Check	OK	04:29	

Data Download Date/Time: 01/28/2016 22:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLEMONS,JASON,R
BREATH TEST OPERATOR AGENCY: FHP Troop K
ARRESTING OFFICER: THACKSTON,ERIK
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 12/21/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG402702
DRY GAS EXPIRY DATE: 01/27/2016
VIOLATION CODE: DUI
ARREST TIME: 22:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: CLEMONS

DATE/TIME OF TEST: 01/16/2016 23:24:42

Diagnostics Check	OK	23:17
Air Blank	0.000	23:18
Control Test	0.077	23:18
Air Blank	0.000	23:19
Subject Sample # 1	0.236	23:19
Breath Volume	1.601	23:19
Air Blank	0.000	23:20
Air Blank	0.000	23:22
Subject Sample # 2	0.235	23:22
Breath Volume	1.324	23:22
Air Blank	0.000	23:23
Control Test	0.077	23:24
Air Blank	0.000	23:24
Diagnostics Check	OK	23:24

Data Download Date/Time: 01/28/2016 22:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORRIVEAU, CHAD, R
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: CORRIVEAU, CHAD
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 12/21/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG402702
DRY GAS EXPIRY DATE: 01/27/2016
VIOLATION CODE: DUI
ARREST TIME: 22:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 01/19/2016 23:55:46

Diagnostics Check	OK	23:52	
Air Blank	0.000	23:53	
Control Test	0.078	23:53	
Air Blank	0.000	23:54	
Subject Sample	# 1	23:54	Subject Test Refused
Breath Volume	0.000	23:54	
Air Blank	0.000	23:54	
Control Test	0.078	23:55	
Air Blank	0.000	23:55	
Diagnostics Check	OK	23:55	

Data Download Date/Time: 01/28/2016 22:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,E,C
BREATH TEST OPERATOR AGENCY: Clermont PD
ARRESTING OFFICER: GARRETT,C
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 12/21/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG402702
DRY GAS EXPIRY DATE: 01/27/2016
VIOLATION CODE: DUI
ARREST TIME: 16:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:44

SUBJECT'S LAST NAME: HNATYSCHAK

DATE/TIME OF TEST: 01/21/2016 17:12:42

Diagnostics Check	OK	17:05
Air Blank	0.000	17:05
Control Test	0.079	17:06
Air Blank	0.000	17:06
Subject Sample # 1	0.250	17:07
Breath Volume	1.933	17:07
Air Blank	0.000	17:08
Air Blank	0.000	17:10
Subject Sample # 2	0.248	17:11
Breath Volume	1.855	17:11
Air Blank	0.000	17:11
Control Test	0.076	17:12
Air Blank	0.000	17:12
Diagnostics Check	OK	17:12

Data Download Date/Time: 01/28/2016 22:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,JOSE,M
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: SEMENTO,JONATHAN
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 12/21/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG402702
DRY GAS EXPIRY DATE: 01/27/2016
VIOLATION CODE: DUI
ARREST TIME: 00:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:38

SUBJECT'S LAST NAME: MARTIN

DATE/TIME OF TEST: 01/25/2016 01:42:26

Diagnostics Check	OK	01:33	
Air Blank	0.000	01:34	
Control Test	0.080	01:34	
Air Blank	0.000	01:35	
Subject Sample	# 1	01:38	No Sample Provided
Breath Volume	0.000	01:38	
Air Blank	0.000	01:38	
Air Blank	0.000	01:40	
Subject Sample	# 2	01:41	Subject Test Refused
Breath Volume	0.000	01:41	
Air Blank	0.000	01:41	
Control Test	0.079	01:41	
Air Blank	0.000	01:42	
Diagnostics Check	OK	01:42	

Data Download Date/Time: 01/28/2016 22:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FOSTER,ANDREW,D
BREATH TEST OPERATOR AGENCY: Leesburg PD
ARRESTING OFFICER: FOSTER,ANDREW
ARRESTING OFFICER AGENCY: Leesburg PD
LAST AGENCY INSPECTION: 12/17/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A1
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:39

SUBJECT'S LAST NAME: DRISCOLL

DATE/TIME OF TEST: 01/01/2016 02:06:07

Diagnostics Check	OK	01:59
Air Blank	0.000	02:00
Control Test	0.078	02:00
Air Blank	0.000	02:01
Subject Sample # 1	0.125	02:01
Breath Volume	2.707	02:01
Air Blank	0.000	02:02
Air Blank	0.000	02:04
Subject Sample # 2	0.119	02:04
Breath Volume	2.753	02:04
Air Blank	0.000	02:05
Control Test	0.079	02:05
Air Blank	0.000	02:06
Diagnostics Check	OK	02:06

Data Download Date/Time: 01/27/2016 14:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRENSHAW,SHANNON,L
BREATH TEST OPERATOR AGENCY: Leesburg PD
ARRESTING OFFICER: JOHNSON,M
ARRESTING OFFICER AGENCY: Leesburg PD
LAST AGENCY INSPECTION: 12/17/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A1
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: ANDRES

DATE/TIME OF TEST: 01/01/2016 02:46:22

Diagnostics Check	OK	02:39
Air Blank	0.000	02:40
Control Test	0.079	02:40
Air Blank	0.000	02:41
Subject Sample # 1	0.087	02:41
Breath Volume	4.121	02:41
Air Blank	0.000	02:42
Air Blank	0.000	02:44
Subject Sample # 2	0.089	02:44
Breath Volume	4.023	02:44
Air Blank	0.000	02:45
Control Test	0.078	02:45
Air Blank	0.000	02:46
Diagnostics Check	OK	02:46

Data Download Date/Time: 01/27/2016 14:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FOSTER,ANDREW,D
BREATH TEST OPERATOR AGENCY: Leesburg PD
ARRESTING OFFICER: JOHNSON,MARK
ARRESTING OFFICER AGENCY: Leesburg PD
LAST AGENCY INSPECTION: 01/27/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A1
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: GRAMS

DATE/TIME OF TEST: 01/29/2016 01:29:28

Diagnostics Check	OK	01:22
Air Blank	0.000	01:22
Control Test	0.079	01:22
Air Blank	0.000	01:23
Subject Sample # 1	0.184	01:24
Breath Volume	2.089	01:24
Air Blank	0.000	01:25
Air Blank	0.000	01:27
Subject Sample # 2	0.168	01:28
Breath Volume	1.699	01:28
Air Blank	0.000	01:28
Control Test	0.078	01:28
Air Blank	0.000	01:29
Diagnostics Check	OK	01:29

Data Download Date/Time: 02/16/2016 04:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEVENS, JEREMY, L
BREATH TEST OPERATOR AGENCY: Leesburg PD
ARRESTING OFFICER: STEVENS, JEREMY
ARRESTING OFFICER AGENCY: Leesburg PD
LAST AGENCY INSPECTION: 01/27/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A1
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: KING

DATE/TIME OF TEST: 01/31/2016 03:52:20

Diagnostics Check	OK	03:45
Air Blank	0.000	03:45
Control Test	0.079	03:46
Air Blank	0.000	03:46
Subject Sample # 1	0.243	03:47
Breath Volume	1.406	03:47
Air Blank	0.000	03:48
Air Blank	0.000	03:49
Subject Sample # 2	0.245	03:50
Breath Volume	1.976	03:50
Air Blank	0.000	03:51
Control Test	0.078	03:51
Air Blank	0.000	03:52
Diagnostics Check	OK	03:52

Data Download Date/Time: 02/16/2016 04:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000825

INSTRUMENT REGISTERED TO: MOUNT DORA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,ANGELA,R
BREATH TEST OPERATOR AGENCY: Mount Dora PD
ARRESTING OFFICER: SMITH,ANGELA
ARRESTING OFFICER AGENCY: Mount Dora PD
LAST AGENCY INSPECTION: 12/05/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1591612
DRY GAS EXPIRY DATE: 11/11/2016
VIOLATION CODE: DUI
ARREST TIME: 17:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:51

SUBJECT'S LAST NAME: DUNCAN

DATE/TIME OF TEST: 01/02/2016 18:20:39

Diagnostics Check	OK	18:17	
Air Blank	0.000	18:17	
Control Test	0.079	18:17	
Air Blank	0.000	18:18	
Subject Sample	# 1	18:19	Subject Test Refused
Breath Volume	0.000	18:19	
Air Blank	0.000	18:19	
Control Test	0.079	18:20	
Air Blank	0.000	18:20	
Diagnostics Check	OK	18:20	

Data Download Date/Time: 01/25/2016 10:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000825

INSTRUMENT REGISTERED TO: MOUNT DORA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,ANGELA,R
BREATH TEST OPERATOR AGENCY: Mount Dora PD
ARRESTING OFFICER: SMITH,ANGELA
ARRESTING OFFICER AGENCY: Mount Dora PD
LAST AGENCY INSPECTION: 12/05/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1591612
DRY GAS EXPIRY DATE: 11/11/2016
VIOLATION CODE: DUI
ARREST TIME: 17:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:51

SUBJECT'S LAST NAME: DUNCAN

DATE/TIME OF TEST: 01/02/2016 19:21:16

Diagnostics Check	OK	19:14
Air Blank	0.000	19:15
Control Test	0.080	19:15
Air Blank	0.000	19:16
Subject Sample # 1	0.137	19:16
Breath Volume	1.972	19:16
Air Blank	0.000	19:17
Air Blank	0.000	19:19
Subject Sample # 2	0.138	19:19
Breath Volume	1.871	19:19
Air Blank	0.000	19:20
Control Test	0.079	19:20
Air Blank	0.000	19:21
Diagnostics Check	OK	19:21

Data Download Date/Time: 01/25/2016 10:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000825

INSTRUMENT REGISTERED TO: MOUNT DORA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,ANGELA,R
BREATH TEST OPERATOR AGENCY: Mount Dora PD
ARRESTING OFFICER: JAEGER,F
ARRESTING OFFICER AGENCY: Mount Dora PD
LAST AGENCY INSPECTION: 12/05/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1591612
DRY GAS EXPIRY DATE: 11/11/2016
VIOLATION CODE: DUI
ARREST TIME: 06:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:16

SUBJECT'S LAST NAME: GEBHARD

DATE/TIME OF TEST: 01/16/2016 07:46:50

Diagnostics Check	OK	07:37
Air Blank	0.000	07:37
Control Test	0.080	07:37
Air Blank	0.000	07:38
Subject Sample # 1	0.121	07:39
Breath Volume	1.804	07:39
Air Blank	0.000	07:39
Air Blank	0.000	07:41
Subject Sample # 2	0.099	07:42
Breath Volume	1.613	07:42
Air Blank	0.000	07:43
Air Blank	0.000	07:44
Subject Sample # 3	0.107	07:45
Breath Volume	1.976	07:45
Air Blank	0.000	07:45
Control Test	0.078	07:46
Air Blank	0.000	07:46
Diagnostics Check	OK	07:46

Data Download Date/Time: 01/25/2016 10:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FRANKLIN,R
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:56

SUBJECT'S LAST NAME: MORFORD

DATE/TIME OF TEST: 01/05/2016 05:25:52

Diagnostics Check	OK	05:19
Air Blank	0.000	05:20
Control Test	0.080	05:20
Air Blank	0.000	05:21
Subject Sample # 1	0.144	05:21
Breath Volume	3.320	05:21
Air Blank	0.000	05:22
Air Blank	0.000	05:23
Subject Sample # 2	0.150	05:24
Breath Volume	3.421	05:24
Air Blank	0.000	05:24
Control Test	0.081	05:25
Air Blank	0.000	05:25
Diagnostics Check	OK	05:25

Data Download Date/Time: 02/01/2016 21:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: GREENE

DATE/TIME OF TEST: 01/06/2016 22:32:10

Diagnostics Check	OK	22:25
Air Blank	0.000	22:26
Control Test	0.080	22:26
Air Blank	0.000	22:26
Subject Sample # 1	0.100	22:27
Breath Volume	2.402	22:27
Air Blank	0.000	22:28
Air Blank	0.000	22:30
Subject Sample # 2	0.101	22:30
Breath Volume	2.515	22:30
Air Blank	0.000	22:31
Control Test	0.079	22:31
Air Blank	0.000	22:32
Diagnostics Check	OK	22:32

Data Download Date/Time: 02/01/2016 21:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DAVIS, M
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:59

SUBJECT'S LAST NAME: AVILES SERRANO

DATE/TIME OF TEST: 01/07/2016 04:29:14

Diagnostics Check	OK	04:21
Air Blank	0.000	04:22
Control Test	0.079	04:22
Air Blank	0.000	04:22
Subject Sample # 1	0.146	04:24
Breath Volume	1.691	04:24
Air Blank	0.000	04:25
Air Blank	0.000	04:27
Subject Sample # 2	0.135	04:27
Breath Volume	1.355	04:27
Air Blank	0.000	04:28
Control Test	0.078	04:28
Air Blank	0.000	04:29
Diagnostics Check	OK	04:29

Data Download Date/Time: 02/01/2016 21:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HERNANDEZ,A
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:57

SUBJECT'S LAST NAME: FUENTES

DATE/TIME OF TEST: 01/08/2016 00:24:19

Diagnostics Check	OK	00:17
Air Blank	0.000	00:18
Control Test	0.080	00:18
Air Blank	0.000	00:18
Subject Sample # 1	0.252	00:19
Breath Volume	1.332	00:19
Air Blank	0.000	00:20
Air Blank	0.000	00:22
Subject Sample # 2	0.260	00:22
Breath Volume	2.496	00:22
Air Blank	0.000	00:23
Control Test	0.079	00:23
Air Blank	0.000	00:24
Diagnostics Check	OK	00:24

Data Download Date/Time: 02/01/2016 21:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WATSON,M
ARRESTING OFFICER AGENCY: U of So FL Campus PD
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:38

SUBJECT'S LAST NAME: SIROTKIN

DATE/TIME OF TEST: 01/08/2016 03:05:09

Diagnostics Check	OK	02:58
Air Blank	0.000	02:59
Control Test	0.080	02:59
Air Blank	0.000	03:00
Subject Sample # 1	0.136	03:00
Breath Volume	2.414	03:00
Air Blank	0.000	03:01
Air Blank	0.000	03:03
Subject Sample # 2	0.133	03:03
Breath Volume	1.554	03:03
Air Blank	0.000	03:04
Control Test	0.078	03:04
Air Blank	0.000	03:05
Diagnostics Check	OK	03:05

Data Download Date/Time: 02/01/2016 21:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DAWSY,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 19:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:29

SUBJECT'S LAST NAME: SHRADER

DATE/TIME OF TEST: 01/08/2016 20:56:26

Diagnostics Check	OK	20:50
Air Blank	0.000	20:50
Control Test	0.079	20:50
Air Blank	0.000	20:51
Subject Sample # 1	0.172	20:51
Breath Volume	2.000	20:51
Air Blank	0.000	20:52
Air Blank	0.000	20:54
Subject Sample # 2	0.175	20:54
Breath Volume	1.683	20:54
Air Blank	0.000	20:55
Control Test	0.079	20:55
Air Blank	0.000	20:56
Diagnostics Check	OK	20:56

Data Download Date/Time: 02/01/2016 21:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK,CASEY,S
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:17

SUBJECT'S LAST NAME: HART

DATE/TIME OF TEST: 01/09/2016 02:46:18

Diagnostics Check	OK	02:39
Air Blank	0.000	02:39
Control Test	0.079	02:40
Air Blank	0.000	02:40
Subject Sample # 1	0.123	02:41
Breath Volume	1.410	02:41
Air Blank	0.000	02:41
Air Blank	0.000	02:43
Subject Sample # 2	0.132	02:44
Breath Volume	2.136	02:44
Air Blank	0.000	02:45
Control Test	0.079	02:45
Air Blank	0.000	02:46
Diagnostics Check	OK	02:46

Data Download Date/Time: 02/01/2016 21:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GOFF,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: COX JR

DATE/TIME OF TEST: 01/09/2016 03:11:34

Diagnostics Check	OK	03:05
Air Blank	0.000	03:05
Control Test	0.079	03:06
Air Blank	0.000	03:06
Subject Sample # 1	0.161	03:07
Breath Volume	2.121	03:07
Air Blank	0.000	03:07
Air Blank	0.000	03:09
Subject Sample # 2	0.153	03:10
Breath Volume	1.777	03:10
Air Blank	0.000	03:10
Control Test	0.079	03:10
Air Blank	0.000	03:11
Diagnostics Check	OK	03:11

Data Download Date/Time: 02/01/2016 21:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BLANCHARD,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:25

SUBJECT'S LAST NAME: PERALES

DATE/TIME OF TEST: 01/09/2016 04:51:25

Diagnostics Check	OK	04:45
Air Blank	0.000	04:45
Control Test	0.078	04:45
Air Blank	0.000	04:46
Subject Sample # 1	0.189	04:46
Breath Volume	1.988	04:46
Air Blank	0.000	04:47
Air Blank	0.000	04:49
Subject Sample # 2	0.193	04:49
Breath Volume	3.179	04:49
Air Blank	0.000	04:50
Control Test	0.079	04:50
Air Blank	0.000	04:51
Diagnostics Check	OK	04:51

Data Download Date/Time: 02/01/2016 21:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY,PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GONZALEZ,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:59

SUBJECT'S LAST NAME: PELUSA

DATE/TIME OF TEST: 01/10/2016 01:30:36

Diagnostics Check	OK	01:20
Air Blank	0.000	01:20
Control Test	0.079	01:20
Air Blank	0.000	01:21
Subject Sample # 1	0.154	01:23
Breath Volume	2.621	01:23
Air Blank	0.000	01:23
Air Blank	0.000	01:25
Subject Sample # 2	0.149	01:29
Breath Volume	1.628	01:29
Air Blank	0.000	01:29
Control Test	0.078	01:30
Air Blank	0.000	01:30
Diagnostics Check	OK	01:30

Data Download Date/Time: 02/01/2016 21:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SARRO,TAYLOR,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:29

SUBJECT'S LAST NAME: DORSEY

DATE/TIME OF TEST: 01/10/2016 02:56:16

Diagnostics Check	OK	02:49
Air Blank	0.000	02:50
Control Test	0.079	02:50
Air Blank	0.000	02:51
Subject Sample # 1	0.037	02:51
Breath Volume	2.628	02:51
Air Blank	0.000	02:52
Air Blank	0.000	02:54
Subject Sample # 2	0.037	02:54
Breath Volume	1.691	02:54
Air Blank	0.000	02:55
Control Test	0.079	02:55
Air Blank	0.000	02:56
Diagnostics Check	OK	02:56

Data Download Date/Time: 02/01/2016 21:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MONTOUTE, S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:24

SUBJECT'S LAST NAME: PAINTER

DATE/TIME OF TEST: 01/10/2016 03:51:22

Diagnostics Check	OK	03:45
Air Blank	0.000	03:45
Control Test	0.079	03:45
Air Blank	0.000	03:46
Subject Sample # 1	0.140	03:46
Breath Volume	3.722	03:46
Air Blank	0.000	03:47
Air Blank	0.000	03:49
Subject Sample # 2	0.147	03:49
Breath Volume	4.253	03:49
Air Blank	0.000	03:50
Control Test	0.079	03:50
Air Blank	0.000	03:51
Diagnostics Check	OK	03:51

Data Download Date/Time: 02/01/2016 21:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SARRO,TAYLOR,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GOFF,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:18

SUBJECT'S LAST NAME: MONTEJO PEREZ

DATE/TIME OF TEST: 01/10/2016 22:45:56

Diagnostics Check	OK	22:39
Air Blank	0.000	22:40
Control Test	0.080	22:40
Air Blank	0.000	22:41
Subject Sample # 1	0.218	22:41
Breath Volume	1.484	22:41
Air Blank	0.000	22:42
Air Blank	0.000	22:44
Subject Sample # 2	0.212	22:44
Breath Volume	1.550	22:44
Air Blank	0.000	22:45
Control Test	0.079	22:45
Air Blank	0.000	22:45
Diagnostics Check	OK	22:45

Data Download Date/Time: 02/01/2016 21:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ANDERSON, C
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:41

SUBJECT'S LAST NAME: MOHAMMED

DATE/TIME OF TEST: 01/11/2016 22:08:12

Diagnostics Check	OK	22:01
Air Blank	0.000	22:02
Control Test	0.080	22:02
Air Blank	0.000	22:02
Subject Sample # 1	0.198	22:03
Breath Volume	1.195	22:03
Air Blank	0.000	22:04
Air Blank	0.000	22:06
Subject Sample # 2	0.196	22:06
Breath Volume	1.761	22:06
Air Blank	0.000	22:07
Control Test	0.080	22:07
Air Blank	0.000	22:08
Diagnostics Check	OK	22:08

Data Download Date/Time: 02/01/2016 21:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SARRO,TAYLOR,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:42

SUBJECT'S LAST NAME: SEYMOUR

DATE/TIME OF TEST: 01/12/2016 00:08:36

Diagnostics Check	OK	00:02
Air Blank	0.000	00:02
Control Test	0.081	00:03
Air Blank	0.000	00:03
Subject Sample # 1	0.097	00:04
Breath Volume	1.945	00:04
Air Blank	0.000	00:04
Air Blank	0.000	00:06
Subject Sample # 2	0.089	00:07
Breath Volume	1.503	00:07
Air Blank	0.000	00:07
Control Test	0.080	00:08
Air Blank	0.000	00:08
Diagnostics Check	OK	00:08

Data Download Date/Time: 02/01/2016 21:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MALONE, C
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:47

SUBJECT'S LAST NAME: STERLING

DATE/TIME OF TEST: 01/12/2016 03:15:22

Diagnostics Check	OK	03:08
Air Blank	0.000	03:08
Control Test	0.080	03:09
Air Blank	0.000	03:09
Subject Sample # 1	0.151	03:10
Breath Volume	3.484	03:10
Air Blank	0.000	03:11
Air Blank	0.000	03:13
Subject Sample # 2	0.153	03:13
Breath Volume	4.300	03:13
Air Blank	0.000	03:14
Control Test	0.081	03:14
Air Blank	0.000	03:15
Diagnostics Check	OK	03:15

Data Download Date/Time: 02/01/2016 21:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: BRONSON

DATE/TIME OF TEST: 01/13/2016 04:22:29

Diagnostics Check	OK	04:16
Air Blank	0.000	04:16
Control Test	0.080	04:17
Air Blank	0.000	04:17
Subject Sample # 1	0.103	04:18
Breath Volume	2.218	04:18
Air Blank	0.000	04:18
Air Blank	0.000	04:20
Subject Sample # 2	0.104	04:20
Breath Volume	2.343	04:20
Air Blank	0.000	04:21
Control Test	0.080	04:21
Air Blank	0.000	04:22
Diagnostics Check	OK	04:22

Data Download Date/Time: 02/01/2016 21:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK,CASEY,S
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VAN-TREESE,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:21

SUBJECT'S LAST NAME: LAUSIER

DATE/TIME OF TEST: 01/13/2016 22:48:33

Diagnostics Check	OK	22:41
Air Blank	0.000	22:41
Control Test	0.081	22:42
Air Blank	0.000	22:42
Subject Sample # 1	0.075	22:43
Breath Volume	1.203	22:43
Air Blank	0.000	22:44
Air Blank	0.000	22:46
Subject Sample # 2	0.076	22:47
Breath Volume	1.253	22:47
Air Blank	0.000	22:47
Control Test	0.080	22:47
Air Blank	0.000	22:48
Diagnostics Check	OK	22:48

Data Download Date/Time: 02/01/2016 21:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARLAUG,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:09

SUBJECT'S LAST NAME: LACHANCE

DATE/TIME OF TEST: 01/14/2016 02:36:17

Diagnostics Check	OK	02:30
Air Blank	0.000	02:30
Control Test	0.080	02:30
Air Blank	0.000	02:31
Subject Sample # 1	0.156	02:31
Breath Volume	1.144	02:31
Air Blank	0.000	02:32
Air Blank	0.000	02:34
Subject Sample # 2	0.164	02:34
Breath Volume	1.492	02:34
Air Blank	0.000	02:35
Control Test	0.080	02:35
Air Blank	0.000	02:36
Diagnostics Check	OK	02:36

Data Download Date/Time: 02/01/2016 21:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARR,BARBARA,H
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PORTMAN,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 15:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:59

SUBJECT'S LAST NAME: BOLTZ

DATE/TIME OF TEST: 01/15/2016 16:26:27

Diagnostics Check	OK	16:20
Air Blank	0.000	16:20
Control Test	0.078	16:21
Air Blank	0.000	16:21
Subject Sample # 1	0.000	16:22
Breath Volume	4.296	16:22
Air Blank	0.000	16:22
Air Blank	0.000	16:24
Subject Sample # 2	0.000	16:25
Breath Volume	4.328	16:25
Air Blank	0.000	16:25
Control Test	0.078	16:25
Air Blank	0.000	16:26
Diagnostics Check	OK	16:26

Data Download Date/Time: 02/01/2016 21:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY,PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MALONE,C
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:12

SUBJECT'S LAST NAME: MURPHY

DATE/TIME OF TEST: 01/16/2016 00:41:28

Diagnostics Check	OK	00:34
Air Blank	0.000	00:35
Control Test	0.080	00:35
Air Blank	0.000	00:35
Subject Sample # 1	0.203	00:36
Breath Volume	2.679	00:36
Air Blank	0.000	00:37
Air Blank	0.000	00:39
Subject Sample # 2	0.204	00:39
Breath Volume	2.796	00:39
Air Blank	0.000	00:40
Control Test	0.079	00:40
Air Blank	0.000	00:41
Diagnostics Check	OK	00:41

Data Download Date/Time: 02/01/2016 21:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MALONE,C
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:21

SUBJECT'S LAST NAME: MORGAN

DATE/TIME OF TEST: 01/16/2016 03:48:06

Diagnostics Check	OK	03:41
Air Blank	0.000	03:41
Control Test	0.079	03:42
Air Blank	0.000	03:42
Subject Sample # 1	0.129	03:43
Breath Volume	2.308	03:43
Air Blank	0.000	03:43
Air Blank	0.000	03:45
Subject Sample # 2	0.144	03:46
Breath Volume	3.785	03:46
Air Blank	0.000	03:47
Control Test	0.079	03:47
Air Blank	0.000	03:48
Diagnostics Check	OK	03:48

Data Download Date/Time: 02/01/2016 21:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GONZALEZ,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:42

SUBJECT'S LAST NAME: ANDREWS

DATE/TIME OF TEST: 01/17/2016 00:09:02

Diagnostics Check	OK	00:02
Air Blank	0.000	00:02
Control Test	0.079	00:03
Air Blank	0.000	00:03
Subject Sample # 1	0.099	00:04
Breath Volume	3.628	00:04
Air Blank	0.000	00:05
Air Blank	0.000	00:07
Subject Sample # 2	0.096	00:07
Breath Volume	1.101	00:07
Air Blank	0.000	00:08
Control Test	0.079	00:08
Air Blank	0.000	00:08
Diagnostics Check	OK	00:09

Data Download Date/Time: 02/01/2016 21:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: EVANS,R
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:01

SUBJECT'S LAST NAME: FLINT

DATE/TIME OF TEST: 01/17/2016 03:28:23

Diagnostics Check	OK	03:21
Air Blank	0.000	03:22
Control Test	0.079	03:22
Air Blank	0.000	03:23
Subject Sample # 1	0.124	03:23
Breath Volume	1.507	03:23
Air Blank	0.000	03:24
Air Blank	0.000	03:26
Subject Sample # 2	0.130	03:26
Breath Volume	3.253	03:26
Air Blank	0.000	03:27
Control Test	0.078	03:27
Air Blank	0.000	03:28
Diagnostics Check	OK	03:28

Data Download Date/Time: 02/01/2016 21:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MALONE,C
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:25

SUBJECT'S LAST NAME: ARCHULETA

DATE/TIME OF TEST: 01/17/2016 05:52:54

Diagnostics Check	OK	05:46
Air Blank	0.000	05:46
Control Test	0.078	05:47
Air Blank	0.000	05:47
Subject Sample # 1	0.104	05:48
Breath Volume	2.484	05:48
Air Blank	0.000	05:48
Air Blank	0.000	05:50
Subject Sample # 2	0.108	05:51
Breath Volume	2.406	05:51
Air Blank	0.000	05:52
Control Test	0.079	05:52
Air Blank	0.000	05:52
Diagnostics Check	OK	05:52

Data Download Date/Time: 02/01/2016 21:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:42

SUBJECT'S LAST NAME: HALLMAN

DATE/TIME OF TEST: 01/18/2016 01:09:01

Diagnostics Check	OK	01:02
Air Blank	0.000	01:02
Control Test	0.080	01:03
Air Blank	0.000	01:03
Subject Sample # 1	0.149	01:04
Breath Volume	1.722	01:04
Air Blank	0.000	01:04
Air Blank	0.000	01:06
Subject Sample # 2	0.156	01:07
Breath Volume	2.171	01:07
Air Blank	0.000	01:08
Control Test	0.079	01:08
Air Blank	0.000	01:08
Diagnostics Check	OK	01:09

Data Download Date/Time: 02/01/2016 21:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: NIEUWENDAM,R
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: RONQUILLO-RIVERA

DATE/TIME OF TEST: 01/18/2016 04:47:02

Diagnostics Check	OK	04:40
Air Blank	0.000	04:41
Control Test	0.080	04:41
Air Blank	0.000	04:41
Subject Sample # 1	0.140	04:42
Breath Volume	2.445	04:42
Air Blank	0.000	04:43
Air Blank	0.000	04:45
Subject Sample # 2	0.139	04:45
Breath Volume	2.843	04:45
Air Blank	0.000	04:46
Control Test	0.080	04:46
Air Blank	0.000	04:46
Diagnostics Check	OK	04:47

Data Download Date/Time: 02/01/2016 21:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DE JESUS,V
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 17:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:15

SUBJECT'S LAST NAME: GILL

DATE/TIME OF TEST: 01/18/2016 19:40:22

Diagnostics Check	OK	19:35	
Air Blank	0.000	19:35	
Control Test	0.081	19:36	
Air Blank	0.000	19:36	
Subject Sample	# 1	19:39	Subject Test Refused
Breath Volume	0.000	19:39	
Air Blank	0.000	19:39	
Control Test	0.081	19:39	
Air Blank	0.000	19:40	
Diagnostics Check	OK	19:40	

Data Download Date/Time: 02/01/2016 21:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:34

SUBJECT'S LAST NAME: O'NEILL

DATE/TIME OF TEST: 01/19/2016 02:01:35

Diagnostics Check	OK	01:54
Air Blank	0.000	01:55
Control Test	0.081	01:55
Air Blank	0.000	01:56
Subject Sample # 1	0.086	01:56
Breath Volume	3.550	01:56
Air Blank	0.000	01:57
Air Blank	0.000	01:59
Subject Sample # 2	0.082	02:00
Breath Volume	3.246	02:00
Air Blank	0.000	02:00
Control Test	0.080	02:01
Air Blank	0.000	02:01
Diagnostics Check	OK	02:01

Data Download Date/Time: 02/01/2016 21:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: CRAGG,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:48

SUBJECT'S LAST NAME: SANCHEZ

DATE/TIME OF TEST: 01/22/2016 01:15:50

Diagnostics Check	OK	01:08
Air Blank	0.000	01:09
Control Test	0.081	01:09
Air Blank	0.000	01:10
Subject Sample # 1	0.058	01:11
Breath Volume	2.507	01:11
Air Blank	0.000	01:12
Air Blank	0.000	01:13
Subject Sample # 2	0.054	01:14
Breath Volume	1.570	01:14
Air Blank	0.000	01:14
Control Test	0.080	01:15
Air Blank	0.000	01:15
Diagnostics Check	OK	01:15

Data Download Date/Time: 02/01/2016 21:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKIPPER,ROGER,G
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARBOSA,F 974/27097
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 07:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:46

SUBJECT'S LAST NAME: SHARP

DATE/TIME OF TEST: 01/22/2016 09:13:43

Diagnostics Check	OK	09:07
Air Blank	0.000	09:07
Control Test	0.080	09:07
Air Blank	0.000	09:08
Subject Sample # 1	0.000	09:09
Breath Volume	1.796	09:09
Air Blank	0.000	09:09
Air Blank	0.000	09:11
Subject Sample # 2	0.000	09:12
Breath Volume	1.398	09:12
Air Blank	0.000	09:12
Control Test	0.079	09:13
Air Blank	0.000	09:13
Diagnostics Check	OK	09:13

Data Download Date/Time: 02/01/2016 21:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MONTOUTE,S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: WILLS JR

DATE/TIME OF TEST: 01/23/2016 03:19:43

Diagnostics Check	OK	03:10
Air Blank	0.000	03:10
Control Test	0.080	03:11
Air Blank	0.000	03:11
Subject Sample # 1	0.195	03:14
Breath Volume	1.859	03:14
Air Blank	0.000	03:15
Air Blank	0.000	03:17
Subject Sample # 2	0.189	03:18
Breath Volume	1.281	03:18
Air Blank	0.000	03:18
Control Test	0.078	03:19
Air Blank	0.000	03:19
Diagnostics Check	OK	03:19

Data Download Date/Time: 02/01/2016 21:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK,CASEY,S
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VAN-TREESE,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:17

SUBJECT'S LAST NAME: MEIGS

DATE/TIME OF TEST: 01/23/2016 04:43:19

Diagnostics Check	OK	04:37
Air Blank	0.000	04:37
Control Test	0.079	04:37
Air Blank	0.000	04:38
Subject Sample # 1	0.122	04:38
Breath Volume	3.687	04:38
Air Blank	0.000	04:39
Air Blank	0.000	04:41
Subject Sample # 2	0.122	04:41
Breath Volume	3.429	04:41
Air Blank	0.000	04:42
Control Test	0.079	04:42
Air Blank	0.000	04:43
Diagnostics Check	OK	04:43

Data Download Date/Time: 02/01/2016 21:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SARRO,TAYLOR,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DAWSY,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 19:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:52

SUBJECT'S LAST NAME: HOPPER

DATE/TIME OF TEST: 01/23/2016 21:20:06

Diagnostics Check	OK	21:14
Air Blank	0.000	21:14
Control Test	0.080	21:14
Air Blank	0.000	21:15
Subject Sample # 1	0.086	21:15
Breath Volume	1.414	21:15
Air Blank	0.000	21:16
Air Blank	0.000	21:18
Subject Sample # 2	0.087	21:18
Breath Volume	1.378	21:18
Air Blank	0.000	21:19
Control Test	0.080	21:19
Air Blank	0.000	21:20
Diagnostics Check	OK	21:20

Data Download Date/Time: 02/01/2016 21:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GOFF,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:07

SUBJECT'S LAST NAME: THORNE

DATE/TIME OF TEST: 01/24/2016 00:34:06

Diagnostics Check	OK	00:27
Air Blank	0.000	00:28
Control Test	0.080	00:28
Air Blank	0.000	00:29
Subject Sample # 1	0.194	00:29
Breath Volume	2.945	00:29
Air Blank	0.000	00:30
Air Blank	0.000	00:32
Subject Sample # 2	0.199	00:32
Breath Volume	2.945	00:32
Air Blank	0.000	00:33
Control Test	0.079	00:33
Air Blank	0.000	00:34
Diagnostics Check	OK	00:34

Data Download Date/Time: 02/01/2016 21:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DE JESUS, V
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:17

SUBJECT'S LAST NAME: WIRTEMBURG

DATE/TIME OF TEST: 01/24/2016 02:44:04

Diagnostics Check	OK	02:37
Air Blank	0.000	02:38
Control Test	0.080	02:38
Air Blank	0.000	02:39
Subject Sample # 1	0.197	02:39
Breath Volume	2.847	02:39
Air Blank	0.000	02:40
Air Blank	0.000	02:42
Subject Sample # 2	0.198	02:42
Breath Volume	2.234	02:42
Air Blank	0.000	02:43
Control Test	0.080	02:43
Air Blank	0.000	02:43
Diagnostics Check	OK	02:44

Data Download Date/Time: 02/01/2016 21:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY, PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNIDER, P
ARRESTING OFFICER AGENCY: Plant City PD
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: RAYO

DATE/TIME OF TEST: 01/24/2016 03:48:34

Diagnostics Check	OK	03:42
Air Blank	0.000	03:42
Control Test	0.081	03:43
Air Blank	0.000	03:43
Subject Sample # 1	0.031	03:44
Breath Volume	2.507	03:44
Air Blank	0.000	03:44
Air Blank	0.000	03:46
Subject Sample # 2	0.031	03:47
Breath Volume	2.566	03:47
Air Blank	0.000	03:47
Control Test	0.081	03:47
Air Blank	0.000	03:48
Diagnostics Check	OK	03:48

Data Download Date/Time: 02/01/2016 21:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MINAYA,J
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:15

SUBJECT'S LAST NAME: WEAVER

DATE/TIME OF TEST: 01/24/2016 05:49:48

Diagnostics Check	OK	05:35	
Air Blank	0.000	05:35	
Control Test	0.081	05:36	
Air Blank	0.000	05:36	
Subject Sample	# 1 0.162	05:39	Volume Not Met
Breath Volume	0.445	05:39	
Air Blank	0.000	05:40	
Air Blank	0.000	05:42	
Subject Sample	# 2 0.167	05:43	
Breath Volume	1.867	05:43	
Air Blank	0.000	05:43	
Air Blank	0.000	05:45	
Subject Sample	# 3 0.160	05:48	
Breath Volume	1.207	05:48	
Air Blank	0.000	05:48	
Control Test	0.080	05:49	
Air Blank	0.000	05:49	
Diagnostics Check	OK	05:49	

Data Download Date/Time: 02/01/2016 21:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SARRO,TAYLOR,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RODRIGUEZ,R
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:07

SUBJECT'S LAST NAME: ARMSTRONG

DATE/TIME OF TEST: 01/24/2016 23:35:06

Diagnostics Check	OK	23:28
Air Blank	0.000	23:29
Control Test	0.080	23:29
Air Blank	0.000	23:29
Subject Sample # 1	0.236	23:30
Breath Volume	3.210	23:30
Air Blank	0.000	23:31
Air Blank	0.000	23:32
Subject Sample # 2	0.232	23:33
Breath Volume	2.582	23:33
Air Blank	0.000	23:34
Control Test	0.080	23:34
Air Blank	0.000	23:35
Diagnostics Check	OK	23:35

Data Download Date/Time: 02/01/2016 21:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY,PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:33

SUBJECT'S LAST NAME: GILBRETH

DATE/TIME OF TEST: 01/25/2016 01:01:11

Diagnostics Check	OK	00:54
Air Blank	0.000	00:55
Control Test	0.081	00:55
Air Blank	0.000	00:56
Subject Sample # 1	0.191	00:56
Breath Volume	2.187	00:56
Air Blank	0.000	00:57
Air Blank	0.000	00:58
Subject Sample # 2	0.207	00:59
Breath Volume	3.648	00:59
Air Blank	0.000	01:00
Control Test	0.081	01:00
Air Blank	0.000	01:01
Diagnostics Check	OK	01:01

Data Download Date/Time: 02/01/2016 21:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SARRO,TAYLOR,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GOFF,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:52

SUBJECT'S LAST NAME: CASEY

DATE/TIME OF TEST: 01/25/2016 01:25:12

Diagnostics Check	OK	01:13	
Air Blank	0.000	01:13	
Control Test	0.080	01:13	
Air Blank	0.000	01:14	
Subject Sample	# 1	01:17	No Sample Provided
Breath Volume	0.000	01:17	
Air Blank	0.000	01:17	
Air Blank	0.000	01:19	
Subject Sample	# 2	0.181	01:20
Breath Volume	1.562	01:20	
Air Blank	0.000	01:21	
Air Blank	0.000	01:23	
Subject Sample	# 3	0.196	01:23
Breath Volume	2.011	01:23	
Air Blank	0.000	01:24	
Control Test	0.080	01:24	
Air Blank	0.000	01:25	
Diagnostics Check	OK	01:25	

Data Download Date/Time: 02/01/2016 21:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK,CASEY,S
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:47

SUBJECT'S LAST NAME: BLAZE

DATE/TIME OF TEST: 01/26/2016 23:15:36

Diagnostics Check	OK	23:08
Air Blank	0.000	23:09
Control Test	0.079	23:09
Air Blank	0.000	23:09
Subject Sample # 1	0.205	23:11
Breath Volume	2.609	23:11
Air Blank	0.000	23:11
Air Blank	0.000	23:13
Subject Sample # 2	0.200	23:14
Breath Volume	1.734	23:14
Air Blank	0.000	23:14
Control Test	0.078	23:15
Air Blank	0.000	23:15
Diagnostics Check	OK	23:15

Data Download Date/Time: 02/01/2016 21:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SEQUEIRA,NC
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:57

SUBJECT'S LAST NAME: SHOATE

DATE/TIME OF TEST: 01/27/2016 00:24:05

Diagnostics Check	OK	00:17
Air Blank	0.000	00:17
Control Test	0.078	00:18
Air Blank	0.000	00:18
Subject Sample # 1	0.202	00:19
Breath Volume	6.109	00:19
Air Blank	0.000	00:19
Air Blank	0.000	00:21
Subject Sample # 2	0.202	00:22
Breath Volume	5.816	00:22
Air Blank	0.000	00:23
Control Test	0.079	00:23
Air Blank	0.000	00:24
Diagnostics Check	OK	00:24

Data Download Date/Time: 02/01/2016 21:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK,CASEY,S
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HERNANDEZ,A
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: AFLOU

DATE/TIME OF TEST: 01/28/2016 02:22:22

Diagnostics Check	OK	02:10	
Air Blank	0.000	02:11	
Control Test	0.080	02:11	
Air Blank	0.000	02:12	
Subject Sample	# 1 0.182	02:15	Volume Not Met
Breath Volume	0.132	02:15	
Air Blank	0.000	02:15	
Air Blank	0.000	02:17	
Subject Sample	# 2 0.176	02:20	Volume Not Met
Breath Volume	0.167	02:20	
Air Blank	0.000	02:21	
Control Test	0.078	02:21	
Air Blank	0.000	02:22	
Diagnostics Check	OK	02:22	

Data Download Date/Time: 02/01/2016 21:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK,CASEY,S
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HERNANDEZ,A
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: AFLOU

DATE/TIME OF TEST: 01/28/2016 02:34:40

Diagnostics Check	OK	02:24	
Air Blank	0.000	02:25	
Control Test	0.078	02:25	
Air Blank	0.000	02:26	
Subject Sample	# 1 0.183	02:29	Volume Not Met
Breath Volume	0.066	02:29	
Air Blank	0.000	02:29	
Air Blank	0.000	02:31	
Subject Sample	# 2	02:33	Subject Test Refused
Breath Volume	0.035	02:33	
Air Blank	0.000	02:33	
Control Test	0.078	02:34	
Air Blank	0.000	02:34	
Diagnostics Check	OK	02:34	

Data Download Date/Time: 02/01/2016 21:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: EHRHART,M
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:28

SUBJECT'S LAST NAME: DYSARD

DATE/TIME OF TEST: 01/28/2016 04:54:46

Diagnostics Check	OK	04:48
Air Blank	0.000	04:48
Control Test	0.079	04:49
Air Blank	0.000	04:49
Subject Sample # 1	0.225	04:50
Breath Volume	3.925	04:50
Air Blank	0.000	04:50
Air Blank	0.000	04:52
Subject Sample # 2	0.229	04:53
Breath Volume	4.281	04:53
Air Blank	0.000	04:53
Control Test	0.079	04:54
Air Blank	0.000	04:54
Diagnostics Check	OK	04:54

Data Download Date/Time: 02/01/2016 21:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SARRO,TAYLOR,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DEJESUS,V
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: LYONS

DATE/TIME OF TEST: 01/29/2016 04:12:02

Diagnostics Check	OK	04:05
Air Blank	0.000	04:05
Control Test	0.080	04:05
Air Blank	0.000	04:06
Subject Sample # 1	0.108	04:07
Breath Volume	3.269	04:07
Air Blank	0.000	04:08
Air Blank	0.000	04:09
Subject Sample # 2	0.108	04:10
Breath Volume	3.644	04:10
Air Blank	0.000	04:11
Control Test	0.079	04:11
Air Blank	0.000	04:11
Diagnostics Check	OK	04:12

Data Download Date/Time: 02/01/2016 21:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SHEA,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: MOREJON

DATE/TIME OF TEST: 01/30/2016 00:26:52

Diagnostics Check	OK	00:20
Air Blank	0.000	00:21
Control Test	0.080	00:21
Air Blank	0.000	00:21
Subject Sample # 1	0.000	00:22
Breath Volume	2.296	00:22
Air Blank	0.000	00:22
Air Blank	0.000	00:24
Subject Sample # 2	0.000	00:25
Breath Volume	1.566	00:25
Air Blank	0.000	00:25
Control Test	0.080	00:26
Air Blank	0.000	00:26
Diagnostics Check	OK	00:26

Data Download Date/Time: 02/01/2016 21:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WOODEN, D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:39

SUBJECT'S LAST NAME: CHARLES

DATE/TIME OF TEST: 01/30/2016 02:11:16

Diagnostics Check	OK	01:59	
Air Blank	0.000	02:00	
Control Test	0.080	02:00	
Air Blank	0.000	02:00	
Subject Sample	# 1 0.131	02:04	Volume Not Met
Breath Volume	0.000	02:04	
Air Blank	0.000	02:04	
Air Blank	0.000	02:06	
Subject Sample	# 2 0.126	02:09	Volume Not Met
Breath Volume	0.937	02:09	
Air Blank	0.000	02:10	
Control Test	0.079	02:10	
Air Blank	0.000	02:11	
Diagnostics Check	OK	02:11	

Data Download Date/Time: 02/01/2016 21:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WOODEN, D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:39

SUBJECT'S LAST NAME: CHARLES

DATE/TIME OF TEST: 01/30/2016 02:20:48

Diagnostics Check	OK	02:12	
Air Blank	0.000	02:12	
Control Test	0.078	02:13	
Air Blank	0.000	02:13	
Subject Sample	# 1 0.134	02:16	Volume Not Met
Breath Volume	0.093	02:16	
Air Blank	0.000	02:17	
Air Blank	0.000	02:19	
Subject Sample	# 2	02:19	Subject Test Refused
Breath Volume	0.000	02:19	
Air Blank	0.000	02:19	
Control Test	0.078	02:20	
Air Blank	0.000	02:20	
Diagnostics Check	OK	02:20	

Data Download Date/Time: 02/01/2016 21:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK,CASEY,S
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: EDGE,S
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:43

SUBJECT'S LAST NAME: MULLINS

DATE/TIME OF TEST: 01/30/2016 23:10:21

Diagnostics Check	OK	23:03
Air Blank	0.000	23:04
Control Test	0.080	23:04
Air Blank	0.000	23:05
Subject Sample # 1	0.168	23:05
Breath Volume	2.984	23:05
Air Blank	0.000	23:06
Air Blank	0.000	23:08
Subject Sample # 2	0.154	23:08
Breath Volume	1.742	23:08
Air Blank	0.000	23:09
Control Test	0.079	23:09
Air Blank	0.000	23:10
Diagnostics Check	OK	23:10

Data Download Date/Time: 02/01/2016 21:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SIMPKINS,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: SIMONE

DATE/TIME OF TEST: 01/31/2016 00:31:52

Diagnostics Check	OK	00:25
Air Blank	0.000	00:26
Control Test	0.080	00:26
Air Blank	0.000	00:26
Subject Sample # 1	0.228	00:27
Breath Volume	2.667	00:27
Air Blank	0.000	00:27
Air Blank	0.000	00:29
Subject Sample # 2	0.219	00:30
Breath Volume	2.394	00:30
Air Blank	0.000	00:30
Control Test	0.079	00:31
Air Blank	0.000	00:31
Diagnostics Check	OK	00:31

Data Download Date/Time: 02/01/2016 21:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: REYNEL

DATE/TIME OF TEST: 01/31/2016 01:28:17

Diagnostics Check	OK	01:21
Air Blank	0.000	01:21
Control Test	0.080	01:22
Air Blank	0.000	01:22
Subject Sample # 1	0.234	01:23
Breath Volume	1.359	01:23
Air Blank	0.000	01:24
Air Blank	0.000	01:26
Subject Sample # 2	0.241	01:26
Breath Volume	1.597	01:26
Air Blank	0.000	01:27
Control Test	0.079	01:27
Air Blank	0.000	01:28
Diagnostics Check	OK	01:28

Data Download Date/Time: 02/01/2016 21:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK,CASEY,S
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: NIEUWENDAM,R
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:07

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 01/31/2016 21:33:50

Diagnostics Check	OK	21:27
Air Blank	0.000	21:27
Control Test	0.080	21:28
Air Blank	0.000	21:28
Subject Sample # 1	0.298	21:29
Breath Volume	1.613	21:29
Air Blank	0.000	21:29
Air Blank	0.000	21:31
Subject Sample # 2	0.283	21:32
Breath Volume	2.261	21:32
Air Blank	0.000	21:32
Control Test	0.079	21:33
Air Blank	0.000	21:33
Diagnostics Check	OK	21:33

Data Download Date/Time: 02/01/2016 21:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 12/10/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12014080A1
DRY GAS EXPIRY DATE: 06/01/2016
VIOLATION CODE: DUI
ARREST TIME: 01:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: WALLACE

DATE/TIME OF TEST: 01/01/2016 02:45:33

Diagnostics Check	OK	02:39
Air Blank	0.000	02:39
Control Test	0.080	02:40
Air Blank	0.000	02:40
Subject Sample # 1	0.127	02:41
Breath Volume	2.433	02:41
Air Blank	0.000	02:41
Air Blank	0.000	02:43
Subject Sample # 2	0.130	02:44
Breath Volume	2.699	02:44
Air Blank	0.000	02:44
Control Test	0.078	02:44
Air Blank	0.000	02:45
Diagnostics Check	OK	02:45

Data Download Date/Time: 01/04/2016 20:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FRANKLIN, R
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/10/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12014080A1
DRY GAS EXPIRY DATE: 06/01/2016
VIOLATION CODE: DUI
ARREST TIME: 04:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:27

SUBJECT'S LAST NAME: WILLIAMSON

DATE/TIME OF TEST: 01/01/2016 05:54:19

Diagnostics Check	OK	05:47
Air Blank	0.000	05:48
Control Test	0.079	05:48
Air Blank	0.000	05:48
Subject Sample # 1	0.120	05:49
Breath Volume	1.273	05:49
Air Blank	0.000	05:49
Air Blank	0.000	05:51
Subject Sample # 2	0.121	05:52
Breath Volume	1.207	05:52
Air Blank	0.000	05:53
Control Test	0.079	05:53
Air Blank	0.000	05:54
Diagnostics Check	OK	05:54

Data Download Date/Time: 01/04/2016 20:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SARRO,TAYLOR,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SHEA,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/10/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12014080A1
DRY GAS EXPIRY DATE: 06/01/2016
VIOLATION CODE: DUI
ARREST TIME: 20:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:38

SUBJECT'S LAST NAME: PASSMAN

DATE/TIME OF TEST: 01/01/2016 21:04:55

Diagnostics Check	OK	20:58
Air Blank	0.000	20:59
Control Test	0.080	20:59
Air Blank	0.000	20:59
Subject Sample # 1	0.112	21:00
Breath Volume	4.632	21:00
Air Blank	0.000	21:00
Air Blank	0.000	21:02
Subject Sample # 2	0.112	21:03
Breath Volume	4.660	21:03
Air Blank	0.000	21:04
Control Test	0.079	21:04
Air Blank	0.000	21:04
Diagnostics Check	OK	21:04

Data Download Date/Time: 01/04/2016 20:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY,PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BLANCHARD,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 12/10/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12014080A1
DRY GAS EXPIRY DATE: 06/01/2016
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:52

SUBJECT'S LAST NAME: HELMLY

DATE/TIME OF TEST: 01/02/2016 02:20:08

Diagnostics Check	OK	02:12
Air Blank	0.000	02:12
Control Test	0.079	02:12
Air Blank	0.000	02:13
Subject Sample # 1	0.173	02:15
Breath Volume	1.660	02:15
Air Blank	0.000	02:15
Air Blank	0.000	02:17
Subject Sample # 2	0.175	02:18
Breath Volume	2.437	02:18
Air Blank	0.000	02:19
Control Test	0.078	02:19
Air Blank	0.000	02:20
Diagnostics Check	OK	02:20

Data Download Date/Time: 01/04/2016 20:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SHEA,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/10/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12014080A1
DRY GAS EXPIRY DATE: 06/01/2016
VIOLATION CODE: DUI
ARREST TIME: 03:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: FARMER

DATE/TIME OF TEST: 01/03/2016 04:21:49

Diagnostics Check	OK	04:15
Air Blank	0.000	04:15
Control Test	0.079	04:16
Air Blank	0.000	04:16
Subject Sample # 1	0.000	04:17
Breath Volume	3.863	04:17
Air Blank	0.000	04:17
Air Blank	0.000	04:19
Subject Sample # 2	0.000	04:20
Breath Volume	2.554	04:20
Air Blank	0.000	04:20
Control Test	0.079	04:21
Air Blank	0.000	04:21
Diagnostics Check	OK	04:21

Data Download Date/Time: 01/04/2016 20:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DAWSY,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12014080A1
DRY GAS EXPIRY DATE: 06/01/2016
VIOLATION CODE: DUI
ARREST TIME: 01:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: HUTSON

DATE/TIME OF TEST: 01/05/2016 02:42:13

Diagnostics Check	OK	02:35
Air Blank	0.000	02:35
Control Test	0.079	02:36
Air Blank	0.000	02:36
Subject Sample # 1	0.128	02:37
Breath Volume	2.382	02:37
Air Blank	0.000	02:38
Air Blank	0.000	02:40
Subject Sample # 2	0.120	02:40
Breath Volume	2.089	02:40
Air Blank	0.000	02:41
Control Test	0.080	02:41
Air Blank	0.000	02:42
Diagnostics Check	OK	02:42

Data Download Date/Time: 02/01/2016 21:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SARRO,TAYLOR,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MALONE,C
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12014080A1
DRY GAS EXPIRY DATE: 06/01/2016
VIOLATION CODE: DUI
ARREST TIME: 02:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:46

SUBJECT'S LAST NAME: PADUA JIMENEZ

DATE/TIME OF TEST: 01/07/2016 03:13:09

Diagnostics Check	OK	03:06
Air Blank	0.000	03:07
Control Test	0.079	03:07
Air Blank	0.000	03:08
Subject Sample # 1	0.000	03:08
Breath Volume	1.917	03:08
Air Blank	0.000	03:09
Air Blank	0.000	03:11
Subject Sample # 2	0.000	03:11
Breath Volume	2.710	03:11
Air Blank	0.000	03:12
Control Test	0.079	03:12
Air Blank	0.000	03:13
Diagnostics Check	OK	03:13

Data Download Date/Time: 02/01/2016 21:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SHEA,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12014080A1
DRY GAS EXPIRY DATE: 06/01/2016
VIOLATION CODE: DUI
ARREST TIME: 03:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:04

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 01/07/2016 04:31:28

Diagnostics Check	OK	04:24
Air Blank	0.000	04:24
Control Test	0.080	04:25
Air Blank	0.000	04:25
Subject Sample # 1	0.140	04:26
Breath Volume	1.343	04:26
Air Blank	0.000	04:26
Air Blank	0.000	04:28
Subject Sample # 2	0.142	04:29
Breath Volume	1.617	04:29
Air Blank	0.000	04:30
Control Test	0.080	04:30
Air Blank	0.000	04:31
Diagnostics Check	OK	04:31

Data Download Date/Time: 02/01/2016 21:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: CRAVERO,A
ARRESTING OFFICER AGENCY: U of So FL Campus PD
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12014080A1
DRY GAS EXPIRY DATE: 06/01/2016
VIOLATION CODE: DUI
ARREST TIME: 01:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:57

SUBJECT'S LAST NAME: ALAM

DATE/TIME OF TEST: 01/08/2016 02:27:18

Diagnostics Check	OK	02:17
Air Blank	0.000	02:18
Control Test	0.079	02:18
Air Blank	0.000	02:18
Subject Sample # 1	0.079	02:21
Breath Volume	1.636	02:21
Air Blank	0.000	02:21
Air Blank	0.000	02:23
Subject Sample # 2	0.071	02:25
Breath Volume	1.496	02:25
Air Blank	0.000	02:26
Control Test	0.079	02:26
Air Blank	0.000	02:27
Diagnostics Check	OK	02:27

Data Download Date/Time: 02/01/2016 21:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK,CASEY,S
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BLANCHARD,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12014080A1
DRY GAS EXPIRY DATE: 06/01/2016
VIOLATION CODE: DUI
ARREST TIME: 22:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:06

SUBJECT'S LAST NAME: FURST

DATE/TIME OF TEST: 01/08/2016 23:32:58

Diagnostics Check	OK	23:26
Air Blank	0.000	23:27
Control Test	0.079	23:27
Air Blank	0.000	23:28
Subject Sample # 1	0.152	23:28
Breath Volume	2.304	23:28
Air Blank	0.000	23:29
Air Blank	0.000	23:30
Subject Sample # 2	0.144	23:31
Breath Volume	1.964	23:31
Air Blank	0.000	23:32
Control Test	0.079	23:32
Air Blank	0.000	23:32
Diagnostics Check	OK	23:32

Data Download Date/Time: 02/01/2016 21:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DE JESUS,V
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12014080A1
DRY GAS EXPIRY DATE: 06/01/2016
VIOLATION CODE: DUI
ARREST TIME: 23:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: BAUTISTA-VELASQUEZ

DATE/TIME OF TEST: 01/09/2016 01:01:51

Diagnostics Check	OK	00:55
Air Blank	0.000	00:55
Control Test	0.079	00:56
Air Blank	0.000	00:56
Subject Sample # 1	0.108	00:57
Breath Volume	2.531	00:57
Air Blank	0.000	00:57
Air Blank	0.000	00:59
Subject Sample # 2	0.112	01:00
Breath Volume	3.613	01:00
Air Blank	0.000	01:00
Control Test	0.079	01:01
Air Blank	0.000	01:01
Diagnostics Check	OK	01:01

Data Download Date/Time: 02/01/2016 21:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK,CASEY,S
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12014080A1
DRY GAS EXPIRY DATE: 06/01/2016
VIOLATION CODE: DUI
ARREST TIME: 02:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:23

SUBJECT'S LAST NAME: WIGGINS-ROYAL

DATE/TIME OF TEST: 01/09/2016 03:58:05

Diagnostics Check	OK	03:45	
Air Blank	0.000	03:46	
Control Test	0.079	03:46	
Air Blank	0.000	03:47	
Subject Sample	# 1 0.193	03:47	
Breath Volume	1.175	03:47	
Air Blank	0.000	03:48	
Air Blank	0.000	03:50	
Subject Sample	# 2 0.181	03:53	Volume Not Met
Breath Volume	0.082	03:53	
Air Blank	0.000	03:54	
Air Blank	0.000	03:55	
Subject Sample	# 3 0.181	03:56	
Breath Volume	1.218	03:56	
Air Blank	0.000	03:57	
Control Test	0.077	03:57	
Air Blank	0.000	03:57	
Diagnostics Check	OK	03:58	

Data Download Date/Time: 02/01/2016 21:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MOSS,J
ARRESTING OFFICER AGENCY: Temple Terrace PD
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12014080A1
DRY GAS EXPIRY DATE: 06/01/2016
VIOLATION CODE: DUI
ARREST TIME: 03:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:59

SUBJECT'S LAST NAME: ANDERSON JR

DATE/TIME OF TEST: 01/09/2016 04:31:47

Diagnostics Check	OK	04:19	
Air Blank	0.000	04:20	
Control Test	0.079	04:20	
Air Blank	0.000	04:20	
Subject Sample	# 1 0.159	04:23	Volume Not Met
Breath Volume	1.023	04:23	
Air Blank	0.000	04:24	
Air Blank	0.000	04:26	
Subject Sample	# 2 0.176	04:27	
Breath Volume	2.097	04:27	
Air Blank	0.000	04:27	
Air Blank	0.000	04:29	
Subject Sample	# 3 0.171	04:30	
Breath Volume	2.140	04:30	
Air Blank	0.000	04:30	
Control Test	0.078	04:31	
Air Blank	0.000	04:31	
Diagnostics Check	OK	04:31	

Data Download Date/Time: 02/01/2016 21:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY, PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LYON, M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12014080A1
DRY GAS EXPIRY DATE: 06/01/2016
VIOLATION CODE: DUI
ARREST TIME: 02:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:59

SUBJECT'S LAST NAME: WAUGH

DATE/TIME OF TEST: 01/10/2016 03:28:02

Diagnostics Check	OK	03:21
Air Blank	0.000	03:22
Control Test	0.080	03:22
Air Blank	0.000	03:23
Subject Sample # 1	0.085	03:23
Breath Volume	2.250	03:23
Air Blank	0.000	03:24
Air Blank	0.000	03:26
Subject Sample # 2	0.092	03:26
Breath Volume	2.164	03:26
Air Blank	0.000	03:27
Control Test	0.080	03:27
Air Blank	0.000	03:27
Diagnostics Check	OK	03:28

Data Download Date/Time: 02/01/2016 21:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRIGHTWELL, R
ARRESTING OFFICER AGENCY: U of So FL Campus PD
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12014080A1
DRY GAS EXPIRY DATE: 06/01/2016
VIOLATION CODE: DUI
ARREST TIME: 04:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:07

SUBJECT'S LAST NAME: CASTILLO

DATE/TIME OF TEST: 01/10/2016 05:34:38

Diagnostics Check	OK	05:27
Air Blank	0.000	05:28
Control Test	0.079	05:28
Air Blank	0.000	05:28
Subject Sample # 1	0.158	05:29
Breath Volume	2.042	05:29
Air Blank	0.000	05:30
Air Blank	0.000	05:32
Subject Sample # 2	0.163	05:33
Breath Volume	1.628	05:33
Air Blank	0.000	05:33
Control Test	0.079	05:34
Air Blank	0.000	05:34
Diagnostics Check	OK	05:34

Data Download Date/Time: 02/01/2016 21:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SARRO,TAYLOR,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LYON,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12014080A1
DRY GAS EXPIRY DATE: 06/01/2016
VIOLATION CODE: DUI
ARREST TIME: 04:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:39

SUBJECT'S LAST NAME: BELCHER

DATE/TIME OF TEST: 01/10/2016 06:06:24

Diagnostics Check	OK	05:59
Air Blank	0.000	06:00
Control Test	0.079	06:00
Air Blank	0.000	06:01
Subject Sample # 1	0.155	06:01
Breath Volume	3.589	06:01
Air Blank	0.000	06:02
Air Blank	0.000	06:04
Subject Sample # 2	0.157	06:04
Breath Volume	2.437	06:04
Air Blank	0.000	06:05
Control Test	0.080	06:05
Air Blank	0.000	06:06
Diagnostics Check	OK	06:06

Data Download Date/Time: 02/01/2016 21:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SARRO,TAYLOR,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: EVANS,R
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12014080A1
DRY GAS EXPIRY DATE: 06/01/2016
VIOLATION CODE: DUI
ARREST TIME: 01:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:28

SUBJECT'S LAST NAME: MORENO MARTIN

DATE/TIME OF TEST: 01/11/2016 02:54:30

Diagnostics Check	OK	02:48
Air Blank	0.000	02:48
Control Test	0.080	02:48
Air Blank	0.000	02:49
Subject Sample # 1	0.125	02:49
Breath Volume	3.312	02:49
Air Blank	0.000	02:50
Air Blank	0.000	02:52
Subject Sample # 2	0.123	02:52
Breath Volume	3.429	02:52
Air Blank	0.000	02:53
Control Test	0.079	02:53
Air Blank	0.000	02:54
Diagnostics Check	OK	02:54

Data Download Date/Time: 02/01/2016 21:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SARRO,TAYLOR,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12014080A1
DRY GAS EXPIRY DATE: 06/01/2016
VIOLATION CODE: DUI
ARREST TIME: 02:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: LEE

DATE/TIME OF TEST: 01/12/2016 03:22:46

Diagnostics Check	OK	03:16
Air Blank	0.000	03:16
Control Test	0.080	03:16
Air Blank	0.000	03:17
Subject Sample # 1	0.240	03:18
Breath Volume	2.312	03:18
Air Blank	0.000	03:18
Air Blank	0.000	03:20
Subject Sample # 2	0.244	03:21
Breath Volume	2.019	03:21
Air Blank	0.000	03:21
Control Test	0.080	03:22
Air Blank	0.000	03:22
Diagnostics Check	OK	03:22

Data Download Date/Time: 02/01/2016 21:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SARRO,TAYLOR,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VAN-TREESE,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12014080A1
DRY GAS EXPIRY DATE: 06/01/2016
VIOLATION CODE: DUI
ARREST TIME: 03:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:22

SUBJECT'S LAST NAME: SIMPSON

DATE/TIME OF TEST: 01/12/2016 03:49:23

Diagnostics Check	OK	03:43
Air Blank	0.000	03:43
Control Test	0.079	03:43
Air Blank	0.000	03:44
Subject Sample # 1	0.099	03:44
Breath Volume	2.218	03:44
Air Blank	0.000	03:45
Air Blank	0.000	03:47
Subject Sample # 2	0.096	03:47
Breath Volume	1.625	03:47
Air Blank	0.000	03:48
Control Test	0.080	03:48
Air Blank	0.000	03:49
Diagnostics Check	OK	03:49

Data Download Date/Time: 02/01/2016 21:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK,CASEY,S
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARLAUG,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12014080A1
DRY GAS EXPIRY DATE: 06/01/2016
VIOLATION CODE: DUI
ARREST TIME: 23:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:07

SUBJECT'S LAST NAME: DONAHOE III

DATE/TIME OF TEST: 01/13/2016 00:34:25

Diagnostics Check	OK	00:28
Air Blank	0.000	00:28
Control Test	0.081	00:28
Air Blank	0.000	00:29
Subject Sample # 1	0.099	00:30
Breath Volume	3.710	00:30
Air Blank	0.000	00:30
Air Blank	0.000	00:32
Subject Sample # 2	0.087	00:32
Breath Volume	1.503	00:32
Air Blank	0.000	00:33
Control Test	0.080	00:33
Air Blank	0.000	00:34
Diagnostics Check	OK	00:34

Data Download Date/Time: 02/01/2016 21:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WATSON,M
ARRESTING OFFICER AGENCY: U of So FL Campus PD
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12014080A1
DRY GAS EXPIRY DATE: 06/01/2016
VIOLATION CODE: DUI
ARREST TIME: 03:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:36

SUBJECT'S LAST NAME: FERRER

DATE/TIME OF TEST: 01/13/2016 05:03:24

Diagnostics Check	OK	04:57
Air Blank	0.000	04:57
Control Test	0.080	04:57
Air Blank	0.000	04:58
Subject Sample # 1	0.179	04:58
Breath Volume	1.214	04:58
Air Blank	0.000	04:59
Air Blank	0.000	05:01
Subject Sample # 2	0.183	05:01
Breath Volume	2.234	05:01
Air Blank	0.000	05:02
Control Test	0.081	05:02
Air Blank	0.000	05:03
Diagnostics Check	OK	05:03

Data Download Date/Time: 02/01/2016 21:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GONZALEZ,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12014080A1
DRY GAS EXPIRY DATE: 06/01/2016
VIOLATION CODE: DUI
ARREST TIME: 21:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:27

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 01/15/2016 22:54:07

Diagnostics Check	OK	22:47
Air Blank	0.000	22:47
Control Test	0.080	22:48
Air Blank	0.000	22:48
Subject Sample # 1	0.163	22:49
Breath Volume	1.179	22:49
Air Blank	0.000	22:49
Air Blank	0.000	22:51
Subject Sample # 2	0.161	22:52
Breath Volume	1.285	22:52
Air Blank	0.000	22:53
Control Test	0.079	22:53
Air Blank	0.000	22:54
Diagnostics Check	OK	22:54

Data Download Date/Time: 02/01/2016 21:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY, PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HERNANDEZ, A
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12014080A1
DRY GAS EXPIRY DATE: 06/01/2016
VIOLATION CODE: DUI
ARREST TIME: 00:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:46

SUBJECT'S LAST NAME: KUST

DATE/TIME OF TEST: 01/16/2016 01:12:53

Diagnostics Check	OK	01:06
Air Blank	0.000	01:06
Control Test	0.079	01:07
Air Blank	0.000	01:07
Subject Sample # 1	0.226	01:08
Breath Volume	3.675	01:08
Air Blank	0.000	01:08
Air Blank	0.000	01:10
Subject Sample # 2	0.228	01:11
Breath Volume	2.199	01:11
Air Blank	0.000	01:11
Control Test	0.079	01:12
Air Blank	0.000	01:12
Diagnostics Check	OK	01:12

Data Download Date/Time: 02/01/2016 21:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SARRO,TAYLOR,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ANGELICCHI,J
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12014080A1
DRY GAS EXPIRY DATE: 06/01/2016
VIOLATION CODE: DUI
ARREST TIME: 01:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:21

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 01/16/2016 02:48:36

Diagnostics Check	OK	02:42
Air Blank	0.000	02:42
Control Test	0.079	02:43
Air Blank	0.000	02:43
Subject Sample # 1	0.182	02:43
Breath Volume	3.445	02:43
Air Blank	0.000	02:44
Air Blank	0.000	02:46
Subject Sample # 2	0.187	02:47
Breath Volume	3.402	02:47
Air Blank	0.000	02:47
Control Test	0.079	02:48
Air Blank	0.000	02:48
Diagnostics Check	OK	02:48

Data Download Date/Time: 02/01/2016 21:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK,CASEY,S
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ANDERSON,C
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12014080A1
DRY GAS EXPIRY DATE: 06/01/2016
VIOLATION CODE: DUI
ARREST TIME: 02:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: GUEVARA MENCHACA

DATE/TIME OF TEST: 01/17/2016 03:21:40

Diagnostics Check	OK	03:15
Air Blank	0.000	03:15
Control Test	0.079	03:16
Air Blank	0.000	03:16
Subject Sample # 1	0.107	03:17
Breath Volume	1.546	03:17
Air Blank	0.000	03:17
Air Blank	0.000	03:19
Subject Sample # 2	0.120	03:20
Breath Volume	2.656	03:20
Air Blank	0.000	03:20
Control Test	0.079	03:21
Air Blank	0.000	03:21
Diagnostics Check	OK	03:21

Data Download Date/Time: 02/01/2016 21:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: EVANS,R
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12014080A1
DRY GAS EXPIRY DATE: 06/01/2016
VIOLATION CODE: DUI
ARREST TIME: 04:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:26

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 01/17/2016 05:53:25

Diagnostics Check	OK	05:47
Air Blank	0.000	05:47
Control Test	0.080	05:47
Air Blank	0.000	05:48
Subject Sample # 1	0.230	05:48
Breath Volume	1.195	05:48
Air Blank	0.000	05:49
Air Blank	0.000	05:51
Subject Sample # 2	0.232	05:51
Breath Volume	1.171	05:51
Air Blank	0.000	05:52
Control Test	0.079	05:52
Air Blank	0.000	05:53
Diagnostics Check	OK	05:53

Data Download Date/Time: 02/01/2016 21:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HERNANDEZ,A
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12014080A1
DRY GAS EXPIRY DATE: 06/01/2016
VIOLATION CODE: DUI
ARREST TIME: 02:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: BROSNAN

DATE/TIME OF TEST: 01/18/2016 03:34:42

Diagnostics Check	OK	03:28
Air Blank	0.000	03:28
Control Test	0.080	03:29
Air Blank	0.000	03:29
Subject Sample # 1	0.185	03:30
Breath Volume	3.421	03:30
Air Blank	0.000	03:30
Air Blank	0.000	03:32
Subject Sample # 2	0.194	03:33
Breath Volume	3.582	03:33
Air Blank	0.000	03:33
Control Test	0.080	03:34
Air Blank	0.000	03:34
Diagnostics Check	OK	03:34

Data Download Date/Time: 02/01/2016 21:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK,CASEY,S
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: NIEUWENDAM,R
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12014080A1
DRY GAS EXPIRY DATE: 06/01/2016
VIOLATION CODE: DUI
ARREST TIME: 03:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:12

SUBJECT'S LAST NAME: ALVARADO

DATE/TIME OF TEST: 01/19/2016 04:39:04

Diagnostics Check	OK	04:32
Air Blank	0.000	04:33
Control Test	0.080	04:33
Air Blank	0.000	04:34
Subject Sample # 1	0.108	04:34
Breath Volume	1.378	04:34
Air Blank	0.000	04:35
Air Blank	0.000	04:37
Subject Sample # 2	0.116	04:37
Breath Volume	1.238	04:37
Air Blank	0.000	04:38
Control Test	0.079	04:38
Air Blank	0.000	04:38
Diagnostics Check	OK	04:39

Data Download Date/Time: 02/01/2016 21:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK,CASEY,S
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12014080A1
DRY GAS EXPIRY DATE: 06/01/2016
VIOLATION CODE: DUI
ARREST TIME: 00:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:22

SUBJECT'S LAST NAME: DAVIS JR

DATE/TIME OF TEST: 01/22/2016 01:49:04

Diagnostics Check	OK	01:42
Air Blank	0.000	01:43
Control Test	0.080	01:43
Air Blank	0.000	01:44
Subject Sample # 1	0.058	01:44
Breath Volume	2.433	01:44
Air Blank	0.000	01:45
Air Blank	0.000	01:47
Subject Sample # 2	0.052	01:47
Breath Volume	1.503	01:47
Air Blank	0.000	01:48
Control Test	0.080	01:48
Air Blank	0.000	01:48
Diagnostics Check	OK	01:49

Data Download Date/Time: 02/01/2016 21:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SHEA,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12014080A1
DRY GAS EXPIRY DATE: 06/01/2016
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:38

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 01/22/2016 03:05:02

Diagnostics Check	OK	02:58
Air Blank	0.000	02:58
Control Test	0.080	02:59
Air Blank	0.000	02:59
Subject Sample # 1	0.222	03:00
Breath Volume	2.000	03:00
Air Blank	0.000	03:01
Air Blank	0.000	03:02
Subject Sample # 2	0.213	03:03
Breath Volume	1.812	03:03
Air Blank	0.000	03:04
Control Test	0.080	03:04
Air Blank	0.000	03:04
Diagnostics Check	OK	03:05

Data Download Date/Time: 02/01/2016 21:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK,CASEY,S
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARLAUG,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12014080A1
DRY GAS EXPIRY DATE: 06/01/2016
VIOLATION CODE: DUI
ARREST TIME: 02:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:44

SUBJECT'S LAST NAME: SANDERS JR

DATE/TIME OF TEST: 01/23/2016 03:10:56

Diagnostics Check	OK	03:04
Air Blank	0.000	03:05
Control Test	0.080	03:05
Air Blank	0.000	03:06
Subject Sample # 1	0.102	03:06
Breath Volume	1.609	03:06
Air Blank	0.000	03:07
Air Blank	0.000	03:09
Subject Sample # 2	0.105	03:09
Breath Volume	1.972	03:09
Air Blank	0.000	03:10
Control Test	0.079	03:10
Air Blank	0.000	03:10
Diagnostics Check	OK	03:10

Data Download Date/Time: 02/01/2016 21:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRABAND,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12014080A1
DRY GAS EXPIRY DATE: 06/01/2016
VIOLATION CODE: DUI
ARREST TIME: 03:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:32

SUBJECT'S LAST NAME: TORRES

DATE/TIME OF TEST: 01/23/2016 03:58:25

Diagnostics Check	OK	03:52
Air Blank	0.000	03:52
Control Test	0.080	03:53
Air Blank	0.000	03:53
Subject Sample # 1	0.068	03:54
Breath Volume	2.445	03:54
Air Blank	0.000	03:54
Air Blank	0.000	03:56
Subject Sample # 2	0.071	03:56
Breath Volume	2.750	03:56
Air Blank	0.000	03:57
Control Test	0.079	03:57
Air Blank	0.000	03:58
Diagnostics Check	OK	03:58

Data Download Date/Time: 02/01/2016 21:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MONTOUTE,S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12014080A1
DRY GAS EXPIRY DATE: 06/01/2016
VIOLATION CODE: DUI
ARREST TIME: 20:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:55

SUBJECT'S LAST NAME: DI GIACOMO

DATE/TIME OF TEST: 01/23/2016 21:21:43

Diagnostics Check	OK	21:15
Air Blank	0.000	21:15
Control Test	0.080	21:16
Air Blank	0.000	21:16
Subject Sample # 1	0.000	21:17
Breath Volume	5.597	21:17
Air Blank	0.000	21:17
Air Blank	0.000	21:19
Subject Sample # 2	0.000	21:20
Breath Volume	4.402	21:20
Air Blank	0.000	21:20
Control Test	0.080	21:21
Air Blank	0.000	21:21
Diagnostics Check	OK	21:21

Data Download Date/Time: 02/01/2016 21:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DAWSY,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12014080A1
DRY GAS EXPIRY DATE: 06/01/2016
VIOLATION CODE: DUI
ARREST TIME: 23:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: FREBURG

DATE/TIME OF TEST: 01/24/2016 01:04:49

Diagnostics Check	OK	01:00	
Air Blank	0.000	01:01	
Control Test	0.081	01:01	
Air Blank	0.000	01:02	
Subject Sample	# 1	01:03	Subject Test Refused
Breath Volume	0.000	01:03	
Air Blank	0.000	01:03	
Control Test	0.080	01:04	
Air Blank	0.000	01:04	
Diagnostics Check	OK	01:04	

Data Download Date/Time: 02/01/2016 21:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY,PATRICIA,D
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: CRESCENZI,T
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12014080A1
DRY GAS EXPIRY DATE: 06/01/2016
VIOLATION CODE: DUI
ARREST TIME: 00:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:46

SUBJECT'S LAST NAME: RUSSELL II

DATE/TIME OF TEST: 01/24/2016 02:12:19

Diagnostics Check	OK	02:06
Air Blank	0.000	02:06
Control Test	0.080	02:06
Air Blank	0.000	02:07
Subject Sample # 1	0.000	02:07
Breath Volume	2.226	02:07
Air Blank	0.000	02:08
Air Blank	0.000	02:10
Subject Sample # 2	0.000	02:10
Breath Volume	1.500	02:10
Air Blank	0.000	02:11
Control Test	0.080	02:11
Air Blank	0.000	02:12
Diagnostics Check	OK	02:12

Data Download Date/Time: 02/01/2016 21:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SARRO,TAYLOR,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FRANKLIN,R
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12014080A1
DRY GAS EXPIRY DATE: 06/01/2016
VIOLATION CODE: DUI
ARREST TIME: 03:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:58

SUBJECT'S LAST NAME: BORJA LEON

DATE/TIME OF TEST: 01/24/2016 04:27:48

Diagnostics Check	OK	04:18
Air Blank	0.000	04:19
Control Test	0.081	04:19
Air Blank	0.000	04:20
Subject Sample # 1	0.178	04:22
Breath Volume	2.054	04:22
Air Blank	0.000	04:23
Air Blank	0.000	04:24
Subject Sample # 2	0.182	04:26
Breath Volume	1.914	04:26
Air Blank	0.000	04:26
Control Test	0.080	04:27
Air Blank	0.000	04:27
Diagnostics Check	OK	04:27

Data Download Date/Time: 02/01/2016 21:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY, PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GONZALEZ, J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12014080A1
DRY GAS EXPIRY DATE: 06/01/2016
VIOLATION CODE: DUI
ARREST TIME: 03:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:16

SUBJECT'S LAST NAME: SILVA-ROMAO

DATE/TIME OF TEST: 01/25/2016 04:43:01

Diagnostics Check	OK	04:36
Air Blank	0.000	04:36
Control Test	0.080	04:37
Air Blank	0.000	04:37
Subject Sample # 1	0.149	04:38
Breath Volume	2.132	04:38
Air Blank	0.000	04:38
Air Blank	0.000	04:40
Subject Sample # 2	0.158	04:41
Breath Volume	2.304	04:41
Air Blank	0.000	04:42
Control Test	0.081	04:42
Air Blank	0.000	04:42
Diagnostics Check	OK	04:43

Data Download Date/Time: 02/01/2016 21:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY, PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ANGELICCHI, J
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12014080A1
DRY GAS EXPIRY DATE: 06/01/2016
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:24

SUBJECT'S LAST NAME: MATEO GARCIA

DATE/TIME OF TEST: 01/26/2016 00:50:38

Diagnostics Check	OK	00:44
Air Blank	0.000	00:44
Control Test	0.080	00:45
Air Blank	0.000	00:45
Subject Sample # 1	0.129	00:46
Breath Volume	2.445	00:46
Air Blank	0.000	00:46
Air Blank	0.000	00:48
Subject Sample # 2	0.132	00:49
Breath Volume	1.562	00:49
Air Blank	0.000	00:49
Control Test	0.080	00:50
Air Blank	0.000	00:50
Diagnostics Check	OK	00:50

Data Download Date/Time: 02/01/2016 21:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12014080A1
DRY GAS EXPIRY DATE: 06/01/2016
VIOLATION CODE: DUI
ARREST TIME: 02:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:02

SUBJECT'S LAST NAME: PARKS

DATE/TIME OF TEST: 01/28/2016 03:30:35

Diagnostics Check	OK	03:22
Air Blank	0.000	03:22
Control Test	0.080	03:23
Air Blank	0.000	03:23
Subject Sample # 1	0.107	03:25
Breath Volume	1.945	03:25
Air Blank	0.000	03:26
Air Blank	0.000	03:28
Subject Sample # 2	0.107	03:29
Breath Volume	2.113	03:29
Air Blank	0.000	03:29
Control Test	0.080	03:29
Air Blank	0.000	03:30
Diagnostics Check	OK	03:30

Data Download Date/Time: 02/01/2016 21:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12014080A1
DRY GAS EXPIRY DATE: 06/01/2016
VIOLATION CODE: DUI
ARREST TIME: 22:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: VAZQUEZ

DATE/TIME OF TEST: 01/29/2016 00:22:39

Diagnostics Check	OK	00:16
Air Blank	0.000	00:16
Control Test	0.080	00:17
Air Blank	0.000	00:17
Subject Sample # 1	0.050	00:18
Breath Volume	2.585	00:18
Air Blank	0.000	00:18
Air Blank	0.000	00:20
Subject Sample # 2	0.051	00:21
Breath Volume	2.175	00:21
Air Blank	0.000	00:21
Control Test	0.080	00:22
Air Blank	0.000	00:22
Diagnostics Check	OK	00:22

Data Download Date/Time: 02/01/2016 21:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SARRO,TAYLOR,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BUCHER,P
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12014080A1
DRY GAS EXPIRY DATE: 06/01/2016
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: CRAIG

DATE/TIME OF TEST: 01/30/2016 02:37:17

Diagnostics Check	OK	02:31
Air Blank	0.000	02:31
Control Test	0.080	02:31
Air Blank	0.000	02:32
Subject Sample # 1	0.112	02:32
Breath Volume	3.042	02:32
Air Blank	0.000	02:33
Air Blank	0.000	02:35
Subject Sample # 2	0.111	02:35
Breath Volume	3.277	02:35
Air Blank	0.000	02:36
Control Test	0.080	02:36
Air Blank	0.000	02:37
Diagnostics Check	OK	02:37

Data Download Date/Time: 02/01/2016 21:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: AGUAYO,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12014080A1
DRY GAS EXPIRY DATE: 06/01/2016
VIOLATION CODE: DUI
ARREST TIME: 23:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:29

SUBJECT'S LAST NAME: ALBINO

DATE/TIME OF TEST: 01/30/2016 23:57:19

Diagnostics Check	OK	23:49
Air Blank	0.000	23:50
Control Test	0.080	23:50
Air Blank	0.000	23:51
Subject Sample # 1	0.141	23:52
Breath Volume	1.777	23:52
Air Blank	0.000	23:52
Air Blank	0.000	23:54
Subject Sample # 2	0.145	23:55
Breath Volume	3.089	23:55
Air Blank	0.000	23:56
Control Test	0.080	23:56
Air Blank	0.000	23:57
Diagnostics Check	OK	23:57

Data Download Date/Time: 02/01/2016 21:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK,CASEY,S
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARLAUG,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12014080A1
DRY GAS EXPIRY DATE: 06/01/2016
VIOLATION CODE: DUI
ARREST TIME: 00:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:03

SUBJECT'S LAST NAME: OREA

DATE/TIME OF TEST: 01/31/2016 01:30:18

Diagnostics Check	OK	01:23
Air Blank	0.000	01:24
Control Test	0.080	01:24
Air Blank	0.000	01:25
Subject Sample # 1	0.203	01:25
Breath Volume	3.199	01:25
Air Blank	0.000	01:26
Air Blank	0.000	01:28
Subject Sample # 2	0.201	01:28
Breath Volume	2.718	01:28
Air Blank	0.000	01:29
Control Test	0.080	01:29
Air Blank	0.000	01:30
Diagnostics Check	OK	01:30

Data Download Date/Time: 02/01/2016 21:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12014080A1
DRY GAS EXPIRY DATE: 06/01/2016
VIOLATION CODE: DUI
ARREST TIME: 03:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:46

SUBJECT'S LAST NAME: RODRIGUEZ JR

DATE/TIME OF TEST: 01/31/2016 04:12:50

Diagnostics Check	OK	04:06
Air Blank	0.000	04:06
Control Test	0.080	04:07
Air Blank	0.000	04:07
Subject Sample # 1	0.108	04:08
Breath Volume	2.164	04:08
Air Blank	0.000	04:08
Air Blank	0.000	04:10
Subject Sample # 2	0.108	04:11
Breath Volume	2.773	04:11
Air Blank	0.000	04:11
Control Test	0.080	04:12
Air Blank	0.000	04:12
Diagnostics Check	OK	04:12

Data Download Date/Time: 02/01/2016 21:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK,CASEY,S
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SHEA,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12014080A1
DRY GAS EXPIRY DATE: 06/01/2016
VIOLATION CODE: DUI
ARREST TIME: 04:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:52

SUBJECT'S LAST NAME: SALGADO JR

DATE/TIME OF TEST: 01/31/2016 05:25:17

Diagnostics Check	OK	05:13	
Air Blank	0.000	05:13	
Control Test	0.080	05:13	
Air Blank	0.000	05:14	
Subject Sample	# 1 0.112	05:17	Volume Not Met
Breath Volume	0.453	05:17	
Air Blank	0.000	05:18	
Air Blank	0.000	05:19	
Subject Sample	# 2 0.140	05:20	
Breath Volume	1.996	05:20	
Air Blank	0.000	05:21	
Air Blank	0.000	05:23	
Subject Sample	# 3 0.138	05:23	
Breath Volume	2.582	05:23	
Air Blank	0.000	05:24	
Control Test	0.079	05:24	
Air Blank	0.000	05:25	
Diagnostics Check	OK	05:25	

Data Download Date/Time: 02/01/2016 21:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUKENS,SHAWN,C
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: LUKENS,SHAWN
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 12/21/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12014080A1
DRY GAS EXPIRY DATE: 06/01/2016
VIOLATION CODE: DUI
ARREST TIME: 18:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:06

SUBJECT'S LAST NAME: MISHOW

DATE/TIME OF TEST: 01/01/2016 18:49:53

Diagnostics Check	OK	18:42
Air Blank	0.000	18:43
Control Test	0.079	18:43
Air Blank	0.000	18:44
Subject Sample # 1	0.193	18:44
Breath Volume	2.429	18:44
Air Blank	0.000	18:45
Air Blank	0.000	18:47
Subject Sample # 2	0.195	18:48
Breath Volume	1.628	18:48
Air Blank	0.000	18:48
Control Test	0.080	18:49
Air Blank	0.000	18:49
Diagnostics Check	OK	18:49

Data Download Date/Time: 01/28/2016 21:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORRIVEAU,CHAD,R
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: CORRIVEAU,CHAD
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 12/21/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12014080A1
DRY GAS EXPIRY DATE: 06/01/2016
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: CHARLES

DATE/TIME OF TEST: 01/08/2016 04:55:17

Diagnostics Check	OK	04:49
Air Blank	0.000	04:49
Control Test	0.079	04:49
Air Blank	0.000	04:50
Subject Sample # 1	0.000	04:50
Breath Volume	1.925	04:50
Air Blank	0.000	04:51
Air Blank	0.000	04:53
Subject Sample # 2	0.000	04:53
Breath Volume	2.437	04:53
Air Blank	0.000	04:54
Control Test	0.079	04:54
Air Blank	0.000	04:55
Diagnostics Check	OK	04:55

Data Download Date/Time: 01/28/2016 21:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHESSHER,SANDRA,W
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: DUNLAP,JASON
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 12/21/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12014080A1
DRY GAS EXPIRY DATE: 06/01/2016
VIOLATION CODE: DUI
ARREST TIME: 02:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:33

SUBJECT'S LAST NAME: ELISCARD

DATE/TIME OF TEST: 01/10/2016 03:56:47

Diagnostics Check	OK	03:53	
Air Blank	0.000	03:54	
Control Test	0.078	03:54	
Air Blank	0.000	03:55	
Subject Sample	# 1	03:55	Subject Test Refused
Breath Volume	0.000	03:55	
Air Blank	0.000	03:55	
Control Test	0.077	03:56	
Air Blank	0.000	03:56	
Diagnostics Check	OK	03:56	

Data Download Date/Time: 01/28/2016 21:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OLDHAM,TAYLOR,D
BREATH TEST OPERATOR AGENCY: Astatula PD
ARRESTING OFFICER: BLOOM,NICHOLAS
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 12/21/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12014080A1
DRY GAS EXPIRY DATE: 06/01/2016
VIOLATION CODE: DUI
ARREST TIME: 03:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: DEPIERRO

DATE/TIME OF TEST: 01/10/2016 04:36:35

Diagnostics Check	OK	04:29
Air Blank	0.000	04:29
Control Test	0.079	04:30
Air Blank	0.000	04:30
Subject Sample # 1	0.199	04:31
Breath Volume	2.230	04:31
Air Blank	0.000	04:32
Air Blank	0.000	04:33
Subject Sample # 2	0.197	04:34
Breath Volume	3.250	04:34
Air Blank	0.000	04:35
Control Test	0.080	04:35
Air Blank	0.000	04:36
Diagnostics Check	OK	04:36

Data Download Date/Time: 01/28/2016 21:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OLDHAM,TAYLOR,D
BREATH TEST OPERATOR AGENCY: Astatula PD
ARRESTING OFFICER: DURYEA,MATTHEW
ARRESTING OFFICER AGENCY: Lady Lake PD
LAST AGENCY INSPECTION: 12/21/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12014080A1
DRY GAS EXPIRY DATE: 06/01/2016
VIOLATION CODE: DUI
ARREST TIME: 18:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:15

SUBJECT'S LAST NAME: MITCHELL

DATE/TIME OF TEST: 01/13/2016 19:46:19

Diagnostics Check	OK	19:38
Air Blank	0.000	19:38
Control Test	0.080	19:39
Air Blank	0.000	19:39
Subject Sample # 1	0.000	19:40
Breath Volume	4.093	19:40
Air Blank	0.000	19:40
Air Blank	0.000	19:42
Subject Sample # 2	0.000	19:44
Breath Volume	3.957	19:44
Air Blank	0.000	19:45
Control Test	0.080	19:45
Air Blank	0.000	19:46
Diagnostics Check	OK	19:46

Data Download Date/Time: 01/28/2016 21:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLLETTE,ERICA,L
BREATH TEST OPERATOR AGENCY: Mount Dora PD
ARRESTING OFFICER: CORTES,VICTOR
ARRESTING OFFICER AGENCY: Eustis PD
LAST AGENCY INSPECTION: 12/21/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12014080A1
DRY GAS EXPIRY DATE: 06/01/2016
VIOLATION CODE: DUI
ARREST TIME: 00:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:18

SUBJECT'S LAST NAME: KRAJNAK

DATE/TIME OF TEST: 01/21/2016 01:41:35

Diagnostics Check	OK	01:38	
Air Blank	0.000	01:39	
Control Test	0.079	01:39	
Air Blank	0.000	01:40	
Subject Sample	# 1	01:40	Subject Test Refused
Breath Volume	0.000	01:40	
Air Blank	0.000	01:40	
Control Test	0.078	01:40	
Air Blank	0.000	01:41	
Diagnostics Check	OK	01:41	

Data Download Date/Time: 01/28/2016 21:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,DAVID,
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: GONZALEZ,DAVID
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 12/21/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12014080A1
DRY GAS EXPIRY DATE: 06/01/2016
VIOLATION CODE: DUI
ARREST TIME: 11:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:06

SUBJECT'S LAST NAME: HEDLUND

DATE/TIME OF TEST: 01/21/2016 02:29:07

Diagnostics Check	OK	02:26	
Air Blank	0.000	02:26	
Control Test	0.079	02:27	
Air Blank	0.000	02:27	
Subject Sample	# 1	02:27	Subject Test Refused
Breath Volume	0.000	02:27	
Air Blank	0.000	02:28	
Control Test	0.079	02:28	
Air Blank	0.000	02:29	
Diagnostics Check	OK	02:29	

Data Download Date/Time: 01/28/2016 21:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHESSHER,SANDRA,W
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: DUNLAP,JASON
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 12/21/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12014080A1
DRY GAS EXPIRY DATE: 06/01/2016
VIOLATION CODE: DUI
ARREST TIME: 22:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:59

SUBJECT'S LAST NAME: WOOD

DATE/TIME OF TEST: 01/22/2016 23:22:51

Diagnostics Check	OK	23:20	
Air Blank	0.000	23:20	
Control Test	0.078	23:20	
Air Blank	0.000	23:21	
Subject Sample	# 1	23:21	Subject Test Refused
Breath Volume	0.000	23:21	
Air Blank	0.000	23:21	
Control Test	0.079	23:22	
Air Blank	0.000	23:22	
Diagnostics Check	OK	23:22	

Data Download Date/Time: 01/28/2016 21:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MIHM,MATTHEW,J
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: MIHM,MATTHEW
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 12/21/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12014080A1
DRY GAS EXPIRY DATE: 06/01/2016
VIOLATION CODE: DUI
ARREST TIME: 22:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:18

SUBJECT'S LAST NAME: SIMMONS

DATE/TIME OF TEST: 01/23/2016 23:47:29

Diagnostics Check	OK	23:46	
Air Blank		23:47	RFI Detect
Air Blank	0.000	23:47	

Data Download Date/Time: 01/28/2016 21:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,DOUG,J
BREATH TEST OPERATOR AGENCY: Tavares PD
ARRESTING OFFICER: DUNLAP,JASON
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 12/21/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12014080A1
DRY GAS EXPIRY DATE: 06/01/2016
VIOLATION CODE: DUI
ARREST TIME: 00:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: JARRELL

DATE/TIME OF TEST: 01/24/2016 01:55:48

Diagnostics Check	OK	01:52	
Air Blank	0.000	01:53	
Control Test	0.079	01:53	
Air Blank	0.000	01:54	
Subject Sample	# 1	01:54	Subject Test Refused
Breath Volume	0.000	01:54	
Air Blank	0.000	01:54	
Control Test	0.079	01:55	
Air Blank	0.000	01:55	
Diagnostics Check	OK	01:55	

Data Download Date/Time: 01/28/2016 21:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,DOUG,J
BREATH TEST OPERATOR AGENCY: Tavares PD
ARRESTING OFFICER: ROBERTS,DOUG
ARRESTING OFFICER AGENCY: Tavares PD
LAST AGENCY INSPECTION: 12/21/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12014080A1
DRY GAS EXPIRY DATE: 06/01/2016
VIOLATION CODE: DUI
ARREST TIME: 19:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:53

SUBJECT'S LAST NAME: RUSSELL

DATE/TIME OF TEST: 01/24/2016 20:22:07

Diagnostics Check	OK	20:18	
Air Blank	0.000	20:19	
Control Test	0.080	20:19	
Air Blank	0.000	20:19	
Subject Sample	# 1	20:20	Subject Test Refused
Breath Volume	0.000	20:20	
Air Blank	0.000	20:21	
Control Test	0.079	20:21	
Air Blank	0.000	20:22	
Diagnostics Check	OK	20:22	

Data Download Date/Time: 01/28/2016 21:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NORGARD,BRYAN,D
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: NORGARD,BRYAN
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 12/21/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12014080A1
DRY GAS EXPIRY DATE: 06/01/2016
VIOLATION CODE: DUI
ARREST TIME: 19:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:02

SUBJECT'S LAST NAME: SHAW

DATE/TIME OF TEST: 01/24/2016 20:39:27

Diagnostics Check	OK	20:26	
Air Blank	0.000	20:26	
Control Test	0.079	20:27	
Air Blank	0.000	20:27	
Subject Sample	# 1 0.105	20:28	
Breath Volume	1.226	20:28	
Air Blank	0.000	20:29	
Air Blank	0.000	20:30	
Subject Sample	# 2 0.101	20:34	Volume Not Met
Breath Volume	0.015	20:34	
Air Blank	0.000	20:34	
Air Blank	0.000	20:36	
Subject Sample	# 3 0.105	20:37	
Breath Volume	1.175	20:37	
Air Blank	0.000	20:38	
Control Test	0.079	20:38	
Air Blank	0.000	20:39	
Diagnostics Check	OK	20:39	

Data Download Date/Time: 01/28/2016 21:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENTON,AMY,B
BREATH TEST OPERATOR AGENCY: Astatula PD
ARRESTING OFFICER: GOEHRING,JONATHAN
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 12/21/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12014080A1
DRY GAS EXPIRY DATE: 06/01/2016
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: DOYLE

DATE/TIME OF TEST: 01/26/2016 02:55:54

Diagnostics Check	OK	02:46
Air Blank	0.000	02:46
Control Test	0.080	02:47
Air Blank	0.000	02:47
Subject Sample # 1	0.201	02:48
Breath Volume	1.671	02:48
Air Blank	0.000	02:49
Air Blank	0.000	02:51
Subject Sample # 2	0.201	02:54
Breath Volume	2.000	02:54
Air Blank	0.000	02:54
Control Test	0.078	02:55
Air Blank	0.000	02:55
Diagnostics Check	OK	02:55

Data Download Date/Time: 01/28/2016 21:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANCE,CHARLES,R
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: LANCE,CHARLES
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 12/14/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1752856
DRY GAS EXPIRY DATE: 10/13/2017
VIOLATION CODE: DUI
ARREST TIME: 03:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: MARTIN

DATE/TIME OF TEST: 01/01/2016 04:30:40

Diagnostics Check	OK	04:24
Air Blank	0.000	04:24
Control Test	0.079	04:24
Air Blank	0.000	04:25
Subject Sample # 1	0.152	04:26
Breath Volume	1.648	04:26
Air Blank	0.000	04:26
Air Blank	0.000	04:28
Subject Sample # 2	0.164	04:29
Breath Volume	2.781	04:29
Air Blank	0.000	04:29
Control Test	0.079	04:30
Air Blank	0.000	04:30
Diagnostics Check	OK	04:30

Data Download Date/Time: 01/27/2016 10:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAYFIELD,MATTHEW,E
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: MAYFIELD,MATTHEW
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 12/14/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1752856
DRY GAS EXPIRY DATE: 10/13/2017
VIOLATION CODE: DUI
ARREST TIME: 11:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 11:35

SUBJECT'S LAST NAME: HUTCHINSON

DATE/TIME OF TEST: 01/01/2016 12:05:59

Diagnostics Check	OK	11:59
Air Blank	0.000	11:59
Control Test	0.079	12:00
Air Blank	0.000	12:00
Subject Sample # 1	0.224	12:01
Breath Volume	2.949	12:01
Air Blank	0.000	12:01
Air Blank	0.000	12:03
Subject Sample # 2	0.224	12:04
Breath Volume	1.371	12:04
Air Blank	0.000	12:05
Control Test	0.080	12:05
Air Blank	0.000	12:05
Diagnostics Check	OK	12:05

Data Download Date/Time: 01/27/2016 10:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANCE,CHARLES,R
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: LANCE,CHARLES
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 12/14/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1752856
DRY GAS EXPIRY DATE: 10/13/2017
VIOLATION CODE: DUI
ARREST TIME: 05:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:15

SUBJECT'S LAST NAME: ATWATER

DATE/TIME OF TEST: 01/12/2016 05:44:51

Diagnostics Check	OK	05:38
Air Blank	0.000	05:39
Control Test	0.080	05:39
Air Blank	0.000	05:39
Subject Sample # 1	0.227	05:40
Breath Volume	1.582	05:40
Air Blank	0.000	05:40
Air Blank	0.000	05:42
Subject Sample # 2	0.231	05:43
Breath Volume	2.308	05:43
Air Blank	0.000	05:43
Control Test	0.079	05:44
Air Blank	0.000	05:44
Diagnostics Check	OK	05:44

Data Download Date/Time: 01/27/2016 10:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,DARYL,R
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: MILLER,DARYL
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 12/14/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1752856
DRY GAS EXPIRY DATE: 10/13/2017
VIOLATION CODE: DUI
ARREST TIME: 15:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:00

SUBJECT'S LAST NAME: MORGAN

DATE/TIME OF TEST: 01/23/2016 16:27:24

Diagnostics Check	OK	16:20
Air Blank	0.000	16:20
Control Test	0.080	16:21
Air Blank	0.000	16:21
Subject Sample # 1	0.326	16:22
Breath Volume	1.488	16:22
Air Blank	0.000	16:23
Air Blank	0.000	16:24
Subject Sample # 2	0.326	16:25
Breath Volume	2.074	16:25
Air Blank	0.000	16:26
Control Test	0.080	16:26
Air Blank	0.000	16:27
Diagnostics Check	OK	16:27

Data Download Date/Time: 01/27/2016 10:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANCE,CHARLES,R
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: LANCE,CHARLES
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 12/14/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1752856
DRY GAS EXPIRY DATE: 10/13/2017
VIOLATION CODE: DUI
ARREST TIME: 00:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: SHAW

DATE/TIME OF TEST: 01/25/2016 00:47:12

Diagnostics Check	OK	00:36
Air Blank	0.000	00:36
Control Test	0.079	00:36
Air Blank	0.000	00:37
Subject Sample # 1	0.142	00:38
Breath Volume	1.480	00:38
Air Blank	0.000	00:38
Air Blank	0.000	00:40
Subject Sample # 2	0.168	00:41
Breath Volume	1.671	00:41
Air Blank	0.000	00:42
Air Blank	0.000	00:43
Subject Sample # 3	0.144	00:45
Breath Volume	1.765	00:45
Air Blank	0.000	00:46
Control Test	0.081	00:46
Air Blank	0.000	00:47
Diagnostics Check	OK	00:47

Data Download Date/Time: 01/27/2016 10:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: LYNN HAVEN PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRELL, CHAD, A
BREATH TEST OPERATOR AGENCY: Lynn Haven PD
ARRESTING OFFICER: HARRELL, CHAD
ARRESTING OFFICER AGENCY: Lynn Haven PD
LAST AGENCY INSPECTION: 12/28/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29614080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 03:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:49

SUBJECT'S LAST NAME: POMPEY

DATE/TIME OF TEST: 01/02/2016 04:45:09

Diagnostics Check	OK	04:33	
Air Blank	0.000	04:34	
Control Test	0.078	04:34	
Air Blank	0.000	04:35	
Subject Sample	# 1	04:38	No Sample Provided
Breath Volume	0.000	04:38	
Air Blank	0.000	04:38	
Air Blank	0.000	04:40	
Subject Sample	# 2	04:43	Subject Test Refused
Breath Volume	0.000	04:43	
Air Blank	0.000	04:44	
Control Test	0.079	04:44	
Air Blank	0.000	04:45	
Diagnostics Check	OK	04:45	

Data Download Date/Time: 01/29/2016 02:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: LYNN HAVEN PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRELL, CHAD, A
BREATH TEST OPERATOR AGENCY: Lynn Haven PD
ARRESTING OFFICER: NEESE, CORBYN
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 12/28/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29614080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 03:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: LAWFORD

DATE/TIME OF TEST: 01/17/2016 04:24:52

Diagnostics Check	OK	04:18
Air Blank	0.000	04:18
Control Test	0.077	04:19
Air Blank	0.000	04:19
Subject Sample # 1	0.274	04:20
Breath Volume	1.742	04:20
Air Blank	0.000	04:21
Air Blank	0.000	04:22
Subject Sample # 2	0.283	04:23
Breath Volume	2.347	04:23
Air Blank	0.000	04:23
Control Test	0.077	04:24
Air Blank	0.000	04:24
Diagnostics Check	OK	04:24

Data Download Date/Time: 01/29/2016 02:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: LYNN HAVEN PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KELLY,JOHN,P
BREATH TEST OPERATOR AGENCY: Lynn Haven PD
ARRESTING OFFICER: THOMPSON,JOSHUA
ARRESTING OFFICER AGENCY: Lynn Haven PD
LAST AGENCY INSPECTION: 12/28/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29614080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 14:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:58

SUBJECT'S LAST NAME: HOSKINS

DATE/TIME OF TEST: 01/17/2016 16:14:44

Diagnostics Check	OK	16:11	
Air Blank	0.000	16:12	
Control Test	0.077	16:12	
Air Blank	0.000	16:13	
Subject Sample	# 1	16:13	Subject Test Refused
Breath Volume	0.000	16:13	
Air Blank	0.000	16:13	
Control Test	0.078	16:14	
Air Blank	0.000	16:14	
Diagnostics Check	OK	16:14	

Data Download Date/Time: 01/29/2016 02:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: LYNN HAVEN PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRELL, CHAD, A
BREATH TEST OPERATOR AGENCY: Lynn Haven PD
ARRESTING OFFICER: BRANCH, JOSH
ARRESTING OFFICER AGENCY: Lynn Haven PD
LAST AGENCY INSPECTION: 12/28/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29614080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 19:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:32

SUBJECT'S LAST NAME: BAILEY

DATE/TIME OF TEST: 01/17/2016 20:07:41

Diagnostics Check	OK	20:04	
Air Blank	0.000	20:05	
Control Test	0.078	20:05	
Air Blank	0.000	20:06	
Subject Sample	# 1	20:06	Subject Test Refused
Breath Volume	0.000	20:06	
Air Blank	0.000	20:06	
Control Test	0.078	20:07	
Air Blank	0.000	20:07	
Diagnostics Check	OK	20:07	

Data Download Date/Time: 01/29/2016 02:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: LYNN HAVEN PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ABRAHAM,ERIC,D
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: ABRAHAM,ERIC
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 01/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29614080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 11:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:02

SUBJECT'S LAST NAME: BATES

DATE/TIME OF TEST: 01/31/2016 11:56:35

Diagnostics Check	OK	11:50
Air Blank	0.000	11:50
Control Test	0.078	11:51
Air Blank	0.000	11:51
Subject Sample # 1	0.000	11:52
Breath Volume	1.972	11:52
Air Blank	0.000	11:52
Air Blank	0.000	11:54
Subject Sample # 2	0.000	11:55
Breath Volume	2.085	11:55
Air Blank	0.000	11:55
Control Test	0.078	11:56
Air Blank	0.000	11:56
Diagnostics Check	OK	11:56

Data Download Date/Time: 02/29/2016 18:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'GRADY,PATRICK,
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: SATTERLEE,ANDREW
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:14

SUBJECT'S LAST NAME: HUBBELL

DATE/TIME OF TEST: 01/01/2016 01:46:39

Diagnostics Check	OK	01:39
Air Blank	0.000	01:39
Control Test	0.080	01:40
Air Blank	0.000	01:40
Subject Sample # 1	0.126	01:41
Breath Volume	1.769	01:41
Air Blank	0.000	01:41
Air Blank	0.000	01:43
Subject Sample # 2	0.127	01:45
Breath Volume	1.300	01:45
Air Blank	0.000	01:45
Control Test	0.079	01:46
Air Blank	0.000	01:46
Diagnostics Check	OK	01:46

Data Download Date/Time: 01/25/2016 10:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'GRADY,PATRICK,
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: O'GRADY,PATRICK
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:23

SUBJECT'S LAST NAME: ACOSTA

DATE/TIME OF TEST: 01/01/2016 01:56:56

Diagnostics Check	OK	01:49
Air Blank	0.000	01:50
Control Test	0.080	01:50
Air Blank	0.000	01:51
Subject Sample # 1	0.262	01:52
Breath Volume	1.230	01:52
Air Blank	0.000	01:52
Air Blank	0.000	01:54
Subject Sample # 2	0.265	01:55
Breath Volume	1.996	01:55
Air Blank	0.000	01:56
Control Test	0.078	01:56
Air Blank	0.000	01:56
Diagnostics Check	OK	01:56

Data Download Date/Time: 01/25/2016 10:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,DANNY,W
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: CRAGIN,STEPHEN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: MEDINA

DATE/TIME OF TEST: 01/01/2016 02:59:05

Diagnostics Check	OK	02:53
Air Blank	0.000	02:53
Control Test	0.079	02:53
Air Blank	0.000	02:54
Subject Sample # 1	0.197	02:54
Breath Volume	1.914	02:54
Air Blank	0.000	02:55
Air Blank	0.000	02:57
Subject Sample # 2	0.190	02:57
Breath Volume	1.371	02:57
Air Blank	0.000	02:58
Control Test	0.079	02:58
Air Blank	0.000	02:59
Diagnostics Check	OK	02:59

Data Download Date/Time: 01/25/2016 10:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SATTERLEE, ANDREW, S
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: SATTERLEE, ANDREW
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:07

SUBJECT'S LAST NAME: CHAIFETZ

DATE/TIME OF TEST: 01/01/2016 21:42:11

Diagnostics Check	OK	21:35
Air Blank	0.000	21:35
Control Test	0.080	21:35
Air Blank	0.000	21:36
Subject Sample # 1	0.152	21:37
Breath Volume	3.261	21:37
Air Blank	0.000	21:37
Air Blank	0.000	21:39
Subject Sample # 2	0.148	21:40
Breath Volume	2.433	21:40
Air Blank	0.000	21:41
Control Test	0.080	21:41
Air Blank	0.000	21:42
Diagnostics Check	OK	21:42

Data Download Date/Time: 01/25/2016 10:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARKLEY,JEFFREY,G
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: BARKLEY,JEFFREY
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: OHARA

DATE/TIME OF TEST: 01/02/2016 00:06:07

Diagnostics Check	Fail	00:05	Temperature Regulation Test
Air Blank	0.000	00:06	

Data Download Date/Time: 01/25/2016 10:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARKLEY,JEFFREY,G
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: BARKLEY,JEFFREY
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: OHARA

DATE/TIME OF TEST: 01/02/2016 00:15:50

Diagnostics Check	OK	00:08
Air Blank	0.000	00:09
Control Test	0.080	00:09
Air Blank	0.000	00:10
Subject Sample # 1	0.180	00:10
Breath Volume	1.144	00:10
Air Blank	0.000	00:11
Air Blank	0.000	00:13
Subject Sample # 2	0.179	00:14
Breath Volume	1.835	00:14
Air Blank	0.000	00:14
Control Test	0.080	00:15
Air Blank	0.000	00:15
Diagnostics Check	OK	00:15

Data Download Date/Time: 01/25/2016 10:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,DANNY,W
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: GRAY,DANNY
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:01

SUBJECT'S LAST NAME: CHAPMAN

DATE/TIME OF TEST: 01/02/2016 03:32:40

Diagnostics Check	OK	03:23
Air Blank	0.000	03:23
Control Test	0.079	03:24
Air Blank	0.000	03:24
Subject Sample # 1	0.056	03:27
Breath Volume	2.093	03:27
Air Blank	0.000	03:28
Air Blank	0.000	03:30
Subject Sample # 2	0.054	03:31
Breath Volume	1.257	03:31
Air Blank	0.000	03:31
Control Test	0.079	03:32
Air Blank	0.000	03:32
Diagnostics Check	OK	03:32

Data Download Date/Time: 01/25/2016 10:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ATKINSON, CLINTON, R
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: KANE, JOHN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 09:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:14

SUBJECT'S LAST NAME: LATESSA

DATE/TIME OF TEST: 01/08/2016 10:01:35

Diagnostics Check	OK	09:52
Air Blank	0.000	09:53
Control Test	0.080	09:53
Air Blank	0.000	09:54
Subject Sample # 1	0.245	09:55
Breath Volume	1.316	09:55
Air Blank	0.000	09:56
Air Blank	0.000	09:58
Subject Sample # 2	0.238	09:59
Breath Volume	1.558	09:59
Air Blank	0.000	10:00
Control Test	0.084	10:00
Air Blank	0.000	10:01
Diagnostics Check	OK	10:01

Data Download Date/Time: 01/25/2016 10:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARSON,JAMES,A
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: ALAMPRESE,JOHN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 19:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:55

SUBJECT'S LAST NAME: SCARAFILE

DATE/TIME OF TEST: 01/08/2016 20:23:03

Diagnostics Check	OK	20:16
Air Blank	0.000	20:17
Control Test	0.080	20:17
Air Blank	0.000	20:18
Subject Sample # 1	0.000	20:18
Breath Volume	3.910	20:18
Air Blank	0.000	20:19
Air Blank	0.000	20:21
Subject Sample # 2	0.000	20:21
Breath Volume	1.906	20:21
Air Blank	0.000	20:22
Control Test	0.080	20:22
Air Blank	0.000	20:22
Diagnostics Check	OK	20:23

Data Download Date/Time: 01/25/2016 10:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,DANNY,W
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: GRAY,DANNY
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: MONET

DATE/TIME OF TEST: 01/09/2016 02:13:31

Diagnostics Check	OK	02:06
Air Blank	0.000	02:07
Control Test	0.081	02:07
Air Blank	0.000	02:08
Subject Sample # 1	0.062	02:08
Breath Volume	1.125	02:08
Air Blank	0.000	02:09
Air Blank	0.000	02:11
Subject Sample # 2	0.062	02:12
Breath Volume	2.230	02:12
Air Blank	0.000	02:12
Control Test	0.079	02:12
Air Blank	0.000	02:13
Diagnostics Check	OK	02:13

Data Download Date/Time: 01/25/2016 10:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREEN,NICHOLAS,R
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: ALAMPRESE,JOHN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:58

SUBJECT'S LAST NAME: ZIMMERMAN

DATE/TIME OF TEST: 01/09/2016 04:43:34

Diagnostics Check	OK	04:37
Air Blank	0.000	04:37
Control Test	0.080	04:38
Air Blank	0.000	04:38
Subject Sample # 1	0.082	04:39
Breath Volume	2.617	04:39
Air Blank	0.000	04:39
Air Blank	0.000	04:41
Subject Sample # 2	0.077	04:42
Breath Volume	1.453	04:42
Air Blank	0.000	04:42
Control Test	0.079	04:42
Air Blank	0.000	04:43
Diagnostics Check	OK	04:43

Data Download Date/Time: 01/25/2016 10:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESTERLE,HUGH,L
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: THOMAS,DANIEL
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:54

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 01/11/2016 00:21:05

Diagnostics Check	OK	00:14
Air Blank	0.000	00:15
Control Test	0.080	00:15
Air Blank	0.000	00:16
Subject Sample # 1	0.181	00:16
Breath Volume	2.125	00:16
Air Blank	0.000	00:17
Air Blank	0.000	00:19
Subject Sample # 2	0.186	00:19
Breath Volume	2.640	00:19
Air Blank	0.000	00:20
Control Test	0.080	00:20
Air Blank	0.000	00:20
Diagnostics Check	OK	00:21

Data Download Date/Time: 01/25/2016 10:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLS,CHARLES,S
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: BARKLEY,JEFFREY
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:54

SUBJECT'S LAST NAME: BLEICH

DATE/TIME OF TEST: 01/12/2016 21:24:41

Diagnostics Check	OK	21:18
Air Blank	0.000	21:18
Control Test	0.080	21:18
Air Blank	0.000	21:19
Subject Sample # 1	0.231	21:20
Breath Volume	3.160	21:20
Air Blank	0.000	21:20
Air Blank	0.000	21:22
Subject Sample # 2	0.224	21:23
Breath Volume	1.617	21:23
Air Blank	0.000	21:23
Control Test	0.080	21:24
Air Blank	0.000	21:24
Diagnostics Check	OK	21:24

Data Download Date/Time: 01/25/2016 10:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREEN,NICHOLAS,R
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: GREEN,NICHOLAS
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: DIVICO

DATE/TIME OF TEST: 01/13/2016 01:19:43

Diagnostics Check	OK	01:13
Air Blank	0.000	01:14
Control Test	0.081	01:14
Air Blank	0.000	01:15
Subject Sample # 1	0.156	01:15
Breath Volume	1.613	01:15
Air Blank	0.000	01:16
Air Blank	0.000	01:17
Subject Sample # 2	0.154	01:18
Breath Volume	1.800	01:18
Air Blank	0.000	01:18
Control Test	0.081	01:19
Air Blank	0.000	01:19
Diagnostics Check	OK	01:19

Data Download Date/Time: 01/25/2016 10:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,DANNY,W
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: ZARILLO,NICHOLAS
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:08

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 01/16/2016 01:38:12

Diagnostics Check	OK	01:30
Air Blank	0.000	01:31
Control Test	0.078	01:31
Air Blank	0.000	01:31
Subject Sample # 1	0.196	01:33
Breath Volume	4.171	01:33
Air Blank	0.000	01:34
Air Blank	0.000	01:36
Subject Sample # 2	0.184	01:36
Breath Volume	1.105	01:36
Air Blank	0.000	01:37
Control Test	0.077	01:37
Air Blank	0.000	01:38
Diagnostics Check	OK	01:38

Data Download Date/Time: 01/25/2016 10:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARSON,JAMES,A
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: CARSON,JAMES
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:48

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 01/18/2016 02:20:55

Diagnostics Check	OK	02:13
Air Blank	0.000	02:14
Control Test	0.079	02:14
Air Blank	0.000	02:14
Subject Sample # 1	0.083	02:15
Breath Volume	2.042	02:15
Air Blank	0.000	02:16
Air Blank	0.000	02:18
Subject Sample # 2	0.080	02:19
Breath Volume	2.468	02:19
Air Blank	0.000	02:20
Control Test	0.080	02:20
Air Blank	0.000	02:20
Diagnostics Check	OK	02:20

Data Download Date/Time: 01/25/2016 10:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,DANNY,W
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: GRAY,DANNY
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:37

SUBJECT'S LAST NAME: CASTILLO

DATE/TIME OF TEST: 01/21/2016 03:07:09

Diagnostics Check	OK	02:59	
Air Blank	0.000	02:59	
Control Test	0.080	03:00	
Air Blank	0.000	03:00	
Subject Sample	# 1 0.177	03:00	
Breath Volume	1.355	03:00	
Air Blank	0.000	03:01	
Air Blank	0.000	03:03	
Subject Sample	# 2	03:05	Subject Test Refused
Breath Volume	0.000	03:05	
Air Blank	0.000	03:06	
Control Test	0.080	03:06	
Air Blank	0.000	03:07	
Diagnostics Check	OK	03:07	

Data Download Date/Time: 01/25/2016 10:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,DANNY,W
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: WESTERLUND EHLERT,TRENT
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: GARAY

DATE/TIME OF TEST: 01/23/2016 04:09:07

Diagnostics Check	OK	04:02
Air Blank	0.000	04:02
Control Test	0.078	04:03
Air Blank	0.000	04:03
Subject Sample # 1	0.165	04:04
Breath Volume	3.773	04:04
Air Blank	0.000	04:05
Air Blank	0.000	04:07
Subject Sample # 2	0.178	04:07
Breath Volume	3.617	04:07
Air Blank	0.000	04:08
Control Test	0.080	04:08
Air Blank	0.000	04:09
Diagnostics Check	OK	04:09

Data Download Date/Time: 01/25/2016 10:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,DANNY,W
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: GREEN,NICHOLAS
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:46

SUBJECT'S LAST NAME: CARL

DATE/TIME OF TEST: 01/23/2016 04:19:46

Diagnostics Check	OK	04:12
Air Blank	0.000	04:13
Control Test	0.080	04:13
Air Blank	0.000	04:14
Subject Sample # 1	0.211	04:14
Breath Volume	2.089	04:14
Air Blank	0.000	04:15
Air Blank	0.000	04:16
Subject Sample # 2	0.208	04:18
Breath Volume	3.117	04:18
Air Blank	0.000	04:18
Control Test	0.079	04:19
Air Blank	0.000	04:19
Diagnostics Check	OK	04:19

Data Download Date/Time: 01/25/2016 10:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREEN,NICHOLAS,R
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: JOHNSON,SARAH
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:37

SUBJECT'S LAST NAME: BUTHERUS

DATE/TIME OF TEST: 01/24/2016 02:04:35

Diagnostics Check	OK	01:57
Air Blank	0.000	01:58
Control Test	0.080	01:58
Air Blank	0.000	01:58
Subject Sample # 1	0.222	01:59
Breath Volume	1.644	01:59
Air Blank	0.000	02:00
Air Blank	0.000	02:02
Subject Sample # 2	0.214	02:03
Breath Volume	2.867	02:03
Air Blank	0.000	02:03
Control Test	0.079	02:04
Air Blank	0.000	02:04
Diagnostics Check	OK	02:04

Data Download Date/Time: 01/25/2016 10:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUSKEVICH,CURT,E
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: STEMPEL,TYSUN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:40

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 01/24/2016 21:10:14

Diagnostics Check	OK	21:04
Air Blank	0.000	21:04
Control Test	0.080	21:04
Air Blank	0.000	21:05
Subject Sample # 1	0.234	21:05
Breath Volume	2.628	21:05
Air Blank	0.000	21:06
Air Blank	0.000	21:08
Subject Sample # 2	0.230	21:08
Breath Volume	1.464	21:08
Air Blank	0.000	21:09
Control Test	0.080	21:09
Air Blank	0.000	21:10
Diagnostics Check	OK	21:10

Data Download Date/Time: 01/25/2016 10:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUSKEVICH,CURT,E
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: JOHNSON,S
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:56

SUBJECT'S LAST NAME: ROLDAN

DATE/TIME OF TEST: 01/24/2016 23:27:16

Diagnostics Check	OK	23:21
Air Blank	0.000	23:21
Control Test	0.080	23:21
Air Blank	0.000	23:22
Subject Sample # 1	0.224	23:22
Breath Volume	3.363	23:22
Air Blank	0.000	23:23
Air Blank	0.000	23:25
Subject Sample # 2	0.221	23:25
Breath Volume	3.210	23:25
Air Blank	0.000	23:26
Control Test	0.080	23:26
Air Blank	0.000	23:27
Diagnostics Check	OK	23:27

Data Download Date/Time: 01/25/2016 10:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MINGS, DEP. G.,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: MINGS, DEP. G.
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 12/28/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28814080A3
DRY GAS EXPIRY DATE: 11/05/2016
VIOLATION CODE: DUI
ARREST TIME: 00:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 01/10/2016 01:01:50

Diagnostics Check	OK	00:55
Air Blank	0.000	00:56
Control Test	0.079	00:56
Air Blank	0.000	00:57
Subject Sample # 1	0.000	00:57
Breath Volume	2.078	00:57
Air Blank	0.000	00:57
Air Blank	0.000	01:00
Subject Sample # 2	0.000	01:00
Breath Volume	1.816	01:00
Air Blank	0.000	01:00
Control Test	0.079	01:01
Air Blank	0.000	01:01
Diagnostics Check	OK	01:01

Data Download Date/Time: 01/27/2016 09:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MINGS,DEPUTY,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: MINGS,DEPUTY G.
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 12/28/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28814080A3
DRY GAS EXPIRY DATE: 11/05/2016
VIOLATION CODE: DUI
ARREST TIME: 02:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: DE LA CRUZ HERNANDEZ

DATE/TIME OF TEST: 01/24/2016 03:12:14

Diagnostics Check	OK	03:05
Air Blank	0.000	03:05
Control Test	0.081	03:05
Air Blank	0.000	03:06
Subject Sample # 1	0.202	03:07
Breath Volume	1.613	03:07
Air Blank	0.000	03:07
Air Blank	0.000	03:09
Subject Sample # 2	0.202	03:10
Breath Volume	2.195	03:10
Air Blank	0.000	03:11
Control Test	0.081	03:11
Air Blank	0.000	03:12
Diagnostics Check	OK	03:12

Data Download Date/Time: 01/27/2016 09:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000862

INSTRUMENT REGISTERED TO: BLOUNTSTOWN PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEMPLE,DARRYL,D
BREATH TEST OPERATOR AGENCY: Blountstown PD
ARRESTING OFFICER: MCCORVEY,BEAU
ARRESTING OFFICER AGENCY: Calhoun County SO
LAST AGENCY INSPECTION: 12/12/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG425202
DRY GAS EXPIRY DATE: 09/09/2016
VIOLATION CODE: DUI
ARREST TIME: 17:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:01

SUBJECT'S LAST NAME: DEROSIA

DATE/TIME OF TEST: 01/24/2016 18:47:44

Diagnostics Check	OK	18:44	
Air Blank	0.000	18:45	
Control Test	0.079	18:45	
Air Blank	0.000	18:46	
Subject Sample	# 1	18:47	Sequence Aborted
Breath Volume	0.000	18:47	
Air Blank	0.000	18:47	

Data Download Date/Time: 01/29/2016 00:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HILL,THOMAS,C
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: HOPP,SETH
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 12/04/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1737087
DRY GAS EXPIRY DATE: 09/12/2017
VIOLATION CODE: DUI
ARREST TIME: 16:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:33

SUBJECT'S LAST NAME: WATKINS

DATE/TIME OF TEST: 01/02/2016 16:59:11

Diagnostics Check	OK	16:53	
Air Blank	0.000	16:54	
Control Test	0.079	16:54	
Air Blank	0.000	16:55	
Subject Sample	# 1 0.296	16:55	
Breath Volume	2.769	16:55	
Air Blank	0.000	16:56	
Air Blank	0.000	16:58	
Subject Sample	# 2 0.314	16:58	
Breath Volume	3.031	16:58	
Air Blank		16:58	Improper Sample
Air Blank	0.000	16:59	

Data Download Date/Time: 01/13/2016 10:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HILL,THOMAS,C
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: HOPP,SETH
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 12/04/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1737087
DRY GAS EXPIRY DATE: 09/12/2017
VIOLATION CODE: DUI
ARREST TIME: 16:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:33

SUBJECT'S LAST NAME: WATKINS

DATE/TIME OF TEST: 01/02/2016 17:14:01

Diagnostics Check	OK	17:07
Air Blank	0.000	17:08
Control Test	0.079	17:08
Air Blank	0.000	17:09
Subject Sample # 1	0.327	17:09
Breath Volume	3.152	17:09
Air Blank	0.000	17:10
Air Blank	0.000	17:12
Subject Sample # 2	0.313	17:12
Breath Volume	2.871	17:12
Air Blank	0.000	17:13
Control Test	0.079	17:13
Air Blank	0.000	17:13
Diagnostics Check	OK	17:14

Data Download Date/Time: 01/13/2016 10:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,DEANNA,N
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: KILMURRAY,CHRISTOPHER
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 12/04/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1816818
DRY GAS EXPIRY DATE: 02/13/2018
VIOLATION CODE: DUI
ARREST TIME: 03:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:24

SUBJECT'S LAST NAME: STURM

DATE/TIME OF TEST: 01/03/2016 03:49:49

Diagnostics Check	OK	03:47	
Air Blank	0.000	03:47	
Control Test	0.079	03:47	
Air Blank	0.000	03:48	
Subject Sample	# 1	03:48	Subject Test Refused
Breath Volume	0.000	03:48	
Air Blank	0.000	03:48	
Control Test	0.079	03:49	
Air Blank	0.000	03:49	
Diagnostics Check	OK	03:49	

Data Download Date/Time: 01/13/2016 10:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,DEANNA,N
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: KILMURRAY,CHRISTOPHER
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 01/09/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1816818
DRY GAS EXPIRY DATE: 02/13/2018
VIOLATION CODE: DUI
ARREST TIME: 23:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:37

SUBJECT'S LAST NAME: ROSES

DATE/TIME OF TEST: 01/13/2016 00:11:20

Diagnostics Check	OK	23:57	
Air Blank	0.000	23:57	
Control Test	0.079	23:58	
Air Blank	0.000	23:58	
Subject Sample	# 1 0.168	00:00	No .020 agreement
Breath Volume	1.867	00:00	
Air Blank	0.000	00:01	
Air Blank	0.000	00:02	
Subject Sample	# 2 0.139	00:05	Volume Not Met
Breath Volume	0.699	00:05	
Air Blank	0.000	00:06	
Air Blank	0.000	00:08	
Subject Sample	# 3 0.146	00:09	No .020 agreement
Breath Volume	1.546	00:09	
Air Blank	0.000	00:10	
Control Test	0.077	00:10	
Air Blank	0.000	00:11	
Diagnostics Check	OK	00:11	

Data Download Date/Time: 01/13/2016 10:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,DEANNA,N
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: KILMURRAY,CHRISTOPHER
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 01/09/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1816818
DRY GAS EXPIRY DATE: 02/13/2018
VIOLATION CODE: DUI
ARREST TIME: 23:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:37

SUBJECT'S LAST NAME: ROSES

DATE/TIME OF TEST: 01/13/2016 00:27:55

Diagnostics Check	OK	00:14	
Air Blank	0.000	00:14	
Control Test	0.079	00:15	
Air Blank	0.000	00:15	
Subject Sample	# 1 0.180	00:17	
Breath Volume	2.003	00:17	
Air Blank	0.000	00:18	
Air Blank	0.000	00:19	
Subject Sample	# 2 0.178	00:23	Volume Not Met
Breath Volume	0.128	00:23	
Air Blank	0.000	00:23	
Air Blank	0.000	00:25	
Subject Sample	# 3 0.166	00:26	
Breath Volume	1.605	00:26	
Air Blank	0.000	00:27	
Control Test	0.076	00:27	
Air Blank	0.000	00:27	
Diagnostics Check	OK	00:27	

Data Download Date/Time: 01/13/2016 10:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,DEANNA,N
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: MADNICK,JASON
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 01/09/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1816818
DRY GAS EXPIRY DATE: 02/13/2018
VIOLATION CODE: DUI
ARREST TIME: 20:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: RAI

DATE/TIME OF TEST: 01/15/2016 21:23:10

Diagnostics Check	OK	21:20	
Air Blank	0.000	21:20	
Control Test	0.079	21:21	
Air Blank	0.000	21:21	
Subject Sample	# 1	21:21	Subject Test Refused
Breath Volume	0.000	21:21	
Air Blank	0.000	21:22	
Control Test	0.079	21:22	
Air Blank	0.000	21:23	
Diagnostics Check	OK	21:23	

Data Download Date/Time: 02/13/2016 12:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VILLANUEVA, MAURICIO,
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: SANCHEZ, LUIS
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 01/09/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1737087
DRY GAS EXPIRY DATE: 09/12/2017
VIOLATION CODE: DUI
ARREST TIME: 05:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:46

SUBJECT'S LAST NAME: PROVEDA MACHECHA

DATE/TIME OF TEST: 01/18/2016 06:20:06

Diagnostics Check	OK	06:09	
Air Blank	0.000	06:10	
Control Test	0.080	06:10	
Air Blank	0.000	06:10	
Subject Sample	# 1 0.157	06:14	Volume Not Met
Breath Volume	0.074	06:14	
Air Blank	0.000	06:14	
Air Blank	0.000	06:16	
Subject Sample	# 2	06:18	Subject Test Refused
Breath Volume	0.089	06:18	
Air Blank	0.000	06:19	
Control Test	0.078	06:19	
Air Blank	0.000	06:20	
Diagnostics Check	OK	06:20	

Data Download Date/Time: 02/13/2016 12:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARRILLO,JULIO,C
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: CARRILLO,JULIO
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 01/09/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1816818
DRY GAS EXPIRY DATE: 02/13/2018
VIOLATION CODE: DUI
ARREST TIME: 23:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:19

SUBJECT'S LAST NAME: MOONEY

DATE/TIME OF TEST: 01/30/2016 00:42:39

Diagnostics Check	OK	00:39	
Air Blank	0.000	00:40	
Control Test	0.079	00:40	
Air Blank	0.000	00:41	
Subject Sample	# 1	00:41	Subject Test Refused
Breath Volume	0.000	00:41	
Air Blank	0.000	00:41	
Control Test	0.079	00:42	
Air Blank	0.000	00:42	
Diagnostics Check	OK	00:42	

Data Download Date/Time: 02/13/2016 12:14