Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000868

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN-REYES,JUAN,C
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: MARTIN-REYES,JUAN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 12/26/2015

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1669440
DRY GAS EXPIRY DATE: 04/19/2017
VIOLATION CODE: DUI
ARREST TIME: 01:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:52

SUBJECT'S LAST NAME: FOWLKES

DATE/TIME OF TEST: 01/01/2016 02:19:08

Diagnostics Check		OK	02:12
Air Blank		0.000	02:12
Control Test		0.079	02:13
Air Blank		0.000	02:13
Subject Sample	# 1	0.105	02:14
Breath Volume		2.863	02:14
Air Blank		0.000	02:15
Air Blank		0.000	02:16
Subject Sample	# 2	0.093	02:17
Breath Volume		1.855	02:17
Air Blank		0.000	02:18
Control Test		0.078	02:18
Air Blank		0.000	02:19
Diagnostics Check		OK	02:19

Data Download Date/Time: 01/07/2016 13:23

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000868

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LONES,SHAUN,G
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: LONES,SHAUN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 12/26/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1669440
DRY GAS EXPIRY DATE: 04/19/2017
VIOLATION CODE: DUI
ARREST TIME: 02:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 01/02/2016 03:38:21

Diagnostics Check		OK	03:35	
Air Blank		0.000	03:36	
Control Test		0.078	03:36	
Air Blank		0.000	03:36	
Subject Sample	# 1		03:36	Subject Test Refused
Breath Volume		0.000	03:36	
Air Blank		0.000	03:37	
Control Test		0.078	03:37	
Air Blank		0.000	03:38	
Diagnostics Check		OK	03:38	

Data Download Date/Time: 01/07/2016 13:23

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000868

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITEMAN,NICHOLIS,J
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: WHITEMAN,NICHOLIS
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 12/26/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1669440
DRY GAS EXPIRY DATE: 04/19/2017
VIOLATION CODE: DUI
ARREST TIME: 03:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:13

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 01/03/2016 03:40:36

Diagnostics Check		OK	03:34
Air Blank		0.000	03:34
Control Test		0.079	03:35
Air Blank		0.000	03:35
Subject Sample	# 1	0.114	03:36
Breath Volume		4.550	03:36
Air Blank		0.000	03:36
Air Blank		0.000	03:38
Subject Sample	# 2	0.113	03:39
Breath Volume		4.015	03:39
Air Blank		0.000	03:39
Control Test		0.079	03:40
Air Blank		0.000	03:40
Diagnostics Check		OK	03:40

Data Download Date/Time: 01/07/2016 13:23

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000868

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALVAREZ,ORLANDO,
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: ALLEN,JOHN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 12/26/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1669440
DRY GAS EXPIRY DATE: 04/19/2017
VIOLATION CODE: DUI
ARREST TIME: 23:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:54

SUBJECT'S LAST NAME: MORAN

DATE/TIME OF TEST: 01/05/2016 00:27:44

Diagnostics Check		OK	00:15	
Air Blank		0.000	00:15	
Control Test		0.078	00:15	
Air Blank		0.000	00:16	
Subject Sample	# 1		00:19	No Sample Provided
Breath Volume		0.000	00:19	
Air Blank		0.000	00:20	
Air Blank		0.000	00:22	
Subject Sample	# 2	0.187	00:22	
Breath Volume		1.507	00:22	
Air Blank		0.000	00:23	
Air Blank		0.000	00:25	
Subject Sample	# 3	0.188	00:26	
Breath Volume		1.738	00:26	
Air Blank		0.000	00:26	
Control Test		0.078	00:27	
Air Blank		0.000	00:27	
Diagnostics Check		OK	00:27	

Data Download Date/Time: 01/07/2016 13:23

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000868

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELMORE,BENJAMIN,R
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: PLUSSA,KYLE
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 01/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1816818
DRY GAS EXPIRY DATE: 02/13/2018
VIOLATION CODE: BUI
ARREST TIME: 18:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:54

SUBJECT'S LAST NAME: JOHANNESSEN

DATE/TIME OF TEST: 01/07/2016 19:20:30

Diagnostics Check		OK	19:14
Air Blank		0.000	19:14
Control Test		0.080	19:14
Air Blank		0.000	19:15
Subject Sample	# 1	0.000	19:16
Breath Volume		3.632	19:16
Air Blank		0.000	19:16
Air Blank		0.000	19:18
Subject Sample	# 2	0.000	19:19
Breath Volume		3.496	19:19
Air Blank		0.000	19:19
Control Test		0.079	19:19
Air Blank		0.000	19:20
Diagnostics Check		OK	19:20

Data Download Date/Time: 02/11/2016 14:09

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000868

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLANTON,TYLER,A
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: SIMO,ANNETTE
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 01/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1816818
DRY GAS EXPIRY DATE: 02/13/2018
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: WOOLSEY

DATE/TIME OF TEST: 01/26/2016 02:01:17

Diagnostics Check		OK	01:55
Air Blank		0.000	01:55
Control Test		0.081	01:55
Air Blank		0.000	01:56
Subject Sample	# 1	0.203	01:56
Breath Volume		2.496	01:56
Air Blank		0.000	01:57
Air Blank		0.000	01:59
Subject Sample	# 2	0.200	01:59
Breath Volume		2.250	01:59
Air Blank		0.000	02:00
Control Test		0.080	02:00
Air Blank		0.000	02:01
Diagnostics Check		OK	02:01

Data Download Date/Time: 02/11/2016 14:09

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,JESSE,A
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: YOUNG,JESSE
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 12/02/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1816818
DRY GAS EXPIRY DATE: 02/13/2018
VIOLATION CODE: DUI
ARREST TIME: 23:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: KEEFE

DATE/TIME OF TEST: 01/01/2016 00:03:12

Diagnostics Check		OK	00:00	
Air Blank		0.000	00:00	
Control Test		0.081	00:01	
Air Blank		0.000	00:01	
Subject Sample	# 1		00:01	Subject Test Refused
Breath Volume		0.000	00:01	
Air Blank		0.000	00:02	
Control Test		0.080	00:02	
Air Blank		0.000	00:03	
Diagnostics Check		OK	00:03	

Data Download Date/Time: 01/06/2016 06:50

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,JESSE,A
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: YOUNG,JESSE
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 12/02/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1816818
DRY GAS EXPIRY DATE: 02/13/2018
VIOLATION CODE: DUI
ARREST TIME: 04:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:02

SUBJECT'S LAST NAME: KOIDAHL

DATE/TIME OF TEST: 01/03/2016 05:29:04

Diagnostics Check		OK	05:26	
Air Blank		0.000	05:26	
Control Test		0.081	05:27	
Air Blank		0.000	05:27	
Subject Sample	# 1		05:27	Subject Test Refused
Breath Volume		0.000	05:27	
Air Blank		0.000	05:28	
Control Test		0.080	05:28	
Air Blank		0.000	05:28	
Diagnostics Check		OK	05:29	

Data Download Date/Time: 01/06/2016 06:50

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,ALEX,
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: RODRIGUEZ,ALEX
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 01/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1816818
DRY GAS EXPIRY DATE: 02/13/2018
VIOLATION CODE: DUI
ARREST TIME: 03:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: JUELL

DATE/TIME OF TEST: 01/15/2016 04:08:24

Diagnostics Check		OK	04:02
Air Blank		0.000	04:02
Control Test		0.081	04:02
Air Blank		0.000	04:03
Subject Sample	# 1	0.177	04:03
Breath Volume		1.449	04:03
Air Blank		0.000	04:04
Air Blank		0.000	04:06
Subject Sample	# 2	0.172	04:06
Breath Volume		1.472	04:06
Air Blank		0.000	04:07
Control Test		0.080	04:07
Air Blank		0.000	04:08
Diagnostics Check		OK	04:08

Data Download Date/Time: 02/10/2016 14:09

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,JESSE,A
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: YOUNG,JESSE
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 01/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1816818
DRY GAS EXPIRY DATE: 02/13/2018
VIOLATION CODE: DUI
ARREST TIME: 04:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:16

SUBJECT'S LAST NAME: VINSON

DATE/TIME OF TEST: 01/18/2016 05:49:26

Diagnostics Check		OK	05:46	
Air Blank		0.000	05:47	
Control Test		0.083	05:47	
Air Blank		0.000	05:47	
Subject Sample	# 1		05:48	Subject Test Refused
Breath Volume		0.000	05:48	
Air Blank		0.000	05:48	
Control Test		0.082	05:48	
Air Blank		0.000	05:49	
Diagnostics Check		OK	05:49	

Data Download Date/Time: 02/10/2016 14:09

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PERLMAN,GREGG,B
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: PERLMAN,GREGG
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 01/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1816818
DRY GAS EXPIRY DATE: 02/13/2018
VIOLATION CODE: DUI
ARREST TIME: 14:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:00

SUBJECT'S LAST NAME: WELLS II

DATE/TIME OF TEST: 01/20/2016 16:23:23

Diagnostics Check		OK	16:20	
Air Blank		0.000	16:21	
Control Test		0.081	16:21	
Air Blank		0.000	16:21	
Subject Sample	# 1		16:22	Subject Test Refused
Breath Volume		0.000	16:22	
Air Blank		0.000	16:22	
Control Test		0.082	16:22	
Air Blank		0.000	16:23	
Diagnostics Check		OK	16:23	

Data Download Date/Time: 02/10/2016 14:09

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CURRUL,ROBERT,G
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: YOUNG,JESSE
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 01/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1816818
DRY GAS EXPIRY DATE: 02/13/2018
VIOLATION CODE: DUI
ARREST TIME: 20:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:20

SUBJECT'S LAST NAME: CATES

DATE/TIME OF TEST: 01/20/2016 20:45:43

Diagnostics Check		OK	20:41	
Air Blank		0.000	20:42	
Control Test		0.082	20:42	
Air Blank		0.000	20:42	
Subject Sample	# 1		20:44	Interferent Detect
Breath Volume		1.640	20:44	
Air Blank		0.000	20:44	
Control Test		0.082	20:45	
Air Blank		0.000	20:45	
Diagnostics Check		OK	20:45	

Data Download Date/Time: 02/10/2016 14:09

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CURRUL,ROBERT,G
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: YOUNG,JESSE
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 01/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1816818
DRY GAS EXPIRY DATE: 02/13/2018
VIOLATION CODE: DUI
ARREST TIME: 20:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:20

SUBJECT'S LAST NAME: CATES

DATE/TIME OF TEST: 01/20/2016 20:54:29

Diagnostics Check		OK	20:47
Air Blank		0.000	20:48
Control Test		0.082	20:48
Air Blank		0.000	20:49
Subject Sample	# 1	0.139	20:49
Breath Volume		1.207	20:49
Air Blank		0.000	20:50
Air Blank		0.000	20:52
Subject Sample	# 2	0.145	20:52
Breath Volume		1.300	20:52
Air Blank		0.000	20:53
Control Test		0.081	20:53
Air Blank		0.000	20:54
Diagnostics Check		OK	20:54

Data Download Date/Time: 02/10/2016 14:09

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,ALEX,
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: NEGRON,WENDY
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 01/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1816818
DRY GAS EXPIRY DATE: 02/13/2018
VIOLATION CODE: DUI
ARREST TIME: 21:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: GORDON

DATE/TIME OF TEST: 01/22/2016 22:46:32

Diagnostics Check		OK	22:42	
Air Blank		0.000	22:43	
Control Test		0.081	22:43	
Air Blank		0.000	22:44	
Subject Sample	# 1		22:45	Subject Test Refused
Breath Volume		0.000	22:45	
Air Blank		0.000	22:45	
Control Test		0.082	22:45	
Air Blank		0.000	22:46	
Diagnostics Check		OK	22:46	

Data Download Date/Time: 02/10/2016 14:09

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,ALEX,
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: RODRIGUEZ,ALEX
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 01/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1816818
DRY GAS EXPIRY DATE: 02/13/2018
VIOLATION CODE: DUI
ARREST TIME: 02:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:19

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 01/23/2016 03:48:48

Diagnostics Check		OK	03:45	
Air Blank		0.000	03:45	
Control Test		0.081	03:46	
Air Blank		0.000	03:46	
Subject Sample	# 1	0.198	03:47	
Breath Volume		2.804	03:47	
Air Blank			03:47	RFI Detect
Air Blank		0.000	03:47	
Control Test		0.081	03:48	
Air Blank		0.000	03:48	
Diagnostics Check		OK	03:48	

Data Download Date/Time: 02/10/2016 14:09

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,ALEX,
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: RODRIGUEZ,ALEX
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 01/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1816818
DRY GAS EXPIRY DATE: 02/13/2018
VIOLATION CODE: DUI
ARREST TIME: 02:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:19

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 01/23/2016 03:56:28

Diagnostics Check		OK	03:50
Air Blank		0.000	03:50
Control Test		0.080	03:50
Air Blank		0.000	03:51
Subject Sample	# 1	0.176	03:51
Breath Volume		1.515	03:51
Air Blank		0.000	03:52
Air Blank		0.000	03:54
Subject Sample	# 2	0.166	03:54
Breath Volume		1.402	03:54
Air Blank		0.000	03:55
Control Test		0.080	03:55
Air Blank		0.000	03:56
Diagnostics Check		OK	03:56

Data Download Date/Time: 02/10/2016 14:09

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,MARTHA,K
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: FISHER,CLIFFORD
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 01/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1816818
DRY GAS EXPIRY DATE: 02/13/2018
VIOLATION CODE: DUI
ARREST TIME: 20:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:37

SUBJECT'S LAST NAME: RAMER

DATE/TIME OF TEST: 01/25/2016 21:07:19

Diagnostics Check		OK	20:59
Air Blank		0.000	20:59
Control Test		0.082	21:00
Air Blank		0.000	21:00
Subject Sample	# 1	0.249	21:02
Breath Volume		2.257	21:02
Air Blank		0.000	21:03
Air Blank		0.000	21:04
Subject Sample	# 2	0.245	21:05
Breath Volume		2.167	21:05
Air Blank		0.000	21:06
Control Test		0.080	21:06
Air Blank		0.000	21:07
Diagnostics Check		OK	21:07

Data Download Date/Time: 02/10/2016 14:09

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,JESSE,A
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: YOUNG,JESSE
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 01/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1816818
DRY GAS EXPIRY DATE: 02/13/2018
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:24

SUBJECT'S LAST NAME: BAKER

DATE/TIME OF TEST: 01/26/2016 23:54:21

Diagnostics Check		OK	23:47
Air Blank		0.000	23:48
Control Test		0.081	23:48
Air Blank		0.000	23:49
Subject Sample	# 1	0.198	23:49
Breath Volume		2.132	23:49
Air Blank		0.000	23:50
Air Blank		0.000	23:52
Subject Sample	# 2	0.197	23:52
Breath Volume		2.066	23:52
Air Blank		0.000	23:53
Control Test		0.081	23:53
Air Blank		0.000	23:54
Diagnostics Check		OK	23:54

Data Download Date/Time: 02/10/2016 14:09

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,ALEX,
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: RODRIGUEZ,ALEX
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 01/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1816818
DRY GAS EXPIRY DATE: 02/13/2018
VIOLATION CODE: DUI
ARREST TIME: 04:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:11

SUBJECT'S LAST NAME: BONEY

DATE/TIME OF TEST: 01/28/2016 05:38:33

Diagnostics Check		OK	05:35	
Air Blank		0.000	05:36	
Control Test		0.081	05:36	
Air Blank		0.000	05:36	
Subject Sample	# 1		05:37	Subject Test Refused
Breath Volume		0.000	05:37	
Air Blank		0.000	05:37	
Control Test		0.081	05:37	
Air Blank		0.000	05:38	
Diagnostics Check		OK	05:38	

Data Download Date/Time: 02/10/2016 14:09

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,ALEX,
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: RODRIGUEZ,ALEX
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 01/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1816818
DRY GAS EXPIRY DATE: 02/13/2018
VIOLATION CODE: DUI
ARREST TIME: 01:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: KRABBE

DATE/TIME OF TEST: 01/29/2016 02:21:25

Diagnostics Check		OK	02:14
Air Blank		0.000	02:14
Control Test		0.082	02:14
Air Blank		0.000	02:15
Subject Sample	# 1	0.212	02:15
Breath Volume		2.406	02:15
Air Blank		0.000	02:16
Air Blank		0.000	02:18
Subject Sample	# 2	0.200	02:19
Breath Volume		1.507	02:19
Air Blank		0.000	02:20
Control Test		0.080	02:20
Air Blank		0.000	02:21
Diagnostics Check		OK	02:21

Data Download Date/Time: 02/10/2016 14:09

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CURRY,SPENCER,A
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: HOPP,AUSTIN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 01/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1816818
DRY GAS EXPIRY DATE: 02/13/2018
VIOLATION CODE: DUI
ARREST TIME: 14:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:12

SUBJECT'S LAST NAME: STOTTS

DATE/TIME OF TEST: 01/29/2016 16:40:31

Diagnostics Check		OK	16:33
Air Blank		0.000	16:33
Control Test		0.082	16:33
Air Blank		0.000	16:34
Subject Sample	# 1	0.000	16:36
Breath Volume		1.250	16:36
Air Blank		0.000	16:36
Air Blank		0.000	16:38
Subject Sample	# 2	0.000	16:39
Breath Volume		1.316	16:39
Air Blank		0.000	16:39
Control Test		0.082	16:39
Air Blank		0.000	16:40
Diagnostics Check		OK	16:40

Data Download Date/Time: 02/10/2016 14:09

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,DANIEL,A
BREATH TEST OPERATOR AGENCY: Coral Gables PD
ARRESTING OFFICER: SMITH,DANIEL
ARRESTING OFFICER AGENCY: Coral Gables PD
LAST AGENCY INSPECTION: 12/24/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12014080A1
DRY GAS EXPIRY DATE: 06/01/2016
VIOLATION CODE: DUI
ARREST TIME: 12:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:11

SUBJECT'S LAST NAME: FERNANDEZ

DATE/TIME OF TEST: 01/06/2016 12:45:54

Diagnostics Check		OK	12:39
Air Blank		0.000	12:40
Control Test		0.075	12:40
Air Blank		0.000	12:41
Subject Sample	# 1	0.202	12:41
Breath Volume		1.625	12:41
Air Blank		0.000	12:42
Air Blank		0.000	12:43
Subject Sample	# 2	0.204	12:44
Breath Volume		1.703	12:44
Air Blank		0.000	12:45
Control Test		0.080	12:45
Air Blank		0.000	12:45
Diagnostics Check		OK	12:45

Data Download Date/Time: 01/29/2016 12:59

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAHA,KHALED,
BREATH TEST OPERATOR AGENCY: Coral Gables PD
ARRESTING OFFICER: MARDIS,RAIZA
ARRESTING OFFICER AGENCY: Coral Gables PD
LAST AGENCY INSPECTION: 12/24/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12014080A1
DRY GAS EXPIRY DATE: 06/01/2016
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:47

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 01/12/2016 03:15:42

Diagnostics Check		OK	03:07	
Air Blank		0.000	03:07	
Control Test		0.080	03:08	
Air Blank		0.000	03:08	
Subject Sample	# 1		03:11	No Sample Provided
Breath Volume		0.000	03:11	
Air Blank		0.000	03:12	
Air Blank		0.000	03:14	
Subject Sample	# 2		03:14	Subject Test Refused
Breath Volume		0.000	03:14	
Air Blank		0.000	03:14	
Control Test		0.080	03:15	
Air Blank		0.000	03:15	
Diagnostics Check		OK	03:15	

Data Download Date/Time: 01/29/2016 12:59

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAHA,KHALED,
BREATH TEST OPERATOR AGENCY: Coral Gables PD
ARRESTING OFFICER: MARDIS,RAIZA
ARRESTING OFFICER AGENCY: Coral Gables PD
LAST AGENCY INSPECTION: 12/24/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12014080A1
DRY GAS EXPIRY DATE: 06/01/2016
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:47

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 01/12/2016 03:21:40

Diagnostics Check		OK	03:18	
Air Blank		0.000	03:19	
Control Test		0.080	03:19	
Air Blank		0.000	03:20	
Subject Sample	# 1		03:20	Subject Test Refused
Breath Volume		0.000	03:20	
Air Blank		0.000	03:20	
Control Test		0.079	03:21	
Air Blank		0.000	03:21	
Diagnostics Check		OK	03:21	

Data Download Date/Time: 01/29/2016 12:59

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ST AMAND,STEVE,N
BREATH TEST OPERATOR AGENCY: Coral Gables PD
ARRESTING OFFICER: LUKSA,ANDREW
ARRESTING OFFICER AGENCY: Coral Gables PD
LAST AGENCY INSPECTION: 12/24/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12014080A1
DRY GAS EXPIRY DATE: 06/01/2016
VIOLATION CODE: DUI
ARREST TIME: 19:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:10

SUBJECT'S LAST NAME: MALDONADO

DATE/TIME OF TEST: 01/14/2016 20:38:15

Diagnostics Check		OK	20:31
Air Blank		0.000	20:31
Control Test		0.079	20:32
Air Blank		0.000	20:32
Subject Sample	# 1	0.192	20:33
Breath Volume		1.710	20:33
Air Blank		0.000	20:34
Air Blank		0.000	20:36
Subject Sample	# 2	0.189	20:36
Breath Volume		1.558	20:36
Air Blank		0.000	20:37
Control Test		0.080	20:37
Air Blank		0.000	20:38
Diagnostics Check		OK	20:38

Data Download Date/Time: 01/29/2016 12:59

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMSON,ROBERT,H
BREATH TEST OPERATOR AGENCY: Coral Gables PD
ARRESTING OFFICER: PULIDO,EDUARDO
ARRESTING OFFICER AGENCY: Coral Gables PD
LAST AGENCY INSPECTION: 12/24/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12014080A1
DRY GAS EXPIRY DATE: 06/01/2016
VIOLATION CODE: DUI
ARREST TIME: 20:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:38

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 01/15/2016 21:24:29

Diagnostics Check		OK	21:21	
Air Blank		0.000	21:21	
Control Test		0.080	21:22	
Air Blank		0.000	21:22	
Subject Sample	# 1		21:23	Subject Test Refused
Breath Volume		0.000	21:23	
Air Blank		0.000	21:23	
Control Test		0.080	21:23	
Air Blank		0.000	21:24	
Diagnostics Check		OK	21:24	

Data Download Date/Time: 01/29/2016 12:59

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,CRISTINO,
BREATH TEST OPERATOR AGENCY: Coral Gables PD
ARRESTING OFFICER: PEREZ,CRISTINO
ARRESTING OFFICER AGENCY: Coral Gables PD
LAST AGENCY INSPECTION: 12/24/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12014080A1
DRY GAS EXPIRY DATE: 06/01/2016
VIOLATION CODE: DUI
ARREST TIME: 22:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:08

SUBJECT'S LAST NAME: PENA

DATE/TIME OF TEST: 01/15/2016 23:32:48

Diagnostics Check		OK	23:29	
Air Blank		0.000	23:29	
Control Test		0.080	23:30	
Air Blank		0.000	23:30	
Subject Sample	# 1		23:31	Subject Test Refused
Breath Volume		0.000	23:31	
Air Blank		0.000	23:31	
Control Test		0.080	23:32	
Air Blank		0.000	23:32	
Diagnostics Check		OK	23:32	

Data Download Date/Time: 01/29/2016 12:59

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWENS,GARRETT,D
BREATH TEST OPERATOR AGENCY: Coral Gables PD
ARRESTING OFFICER: SOBALVARRO,PAOLA
ARRESTING OFFICER AGENCY: Coral Gables PD
LAST AGENCY INSPECTION: 12/24/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12014080A1
DRY GAS EXPIRY DATE: 06/01/2016
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:46

SUBJECT'S LAST NAME: WAGNER

DATE/TIME OF TEST: 01/16/2016 04:10:38

Diagnostics Check		OK	04:07	
Air Blank		0.000	04:08	
Control Test		0.080	04:08	
Air Blank		0.000	04:09	
Subject Sample	# 1		04:09	Subject Test Refused
Breath Volume		0.000	04:09	
Air Blank		0.000	04:09	
Control Test		0.080	04:10	
Air Blank		0.000	04:10	
Diagnostics Check		OK	04:10	

Data Download Date/Time: 01/29/2016 12:59

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLORES,EMILIO,
BREATH TEST OPERATOR AGENCY: Coral Gables PD
ARRESTING OFFICER: FLORES,EMILIO
ARRESTING OFFICER AGENCY: Coral Gables PD
LAST AGENCY INSPECTION: 12/24/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12014080A1
DRY GAS EXPIRY DATE: 06/01/2016
VIOLATION CODE: DUI
ARREST TIME: 09:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:37

SUBJECT'S LAST NAME: MONTENEGRO LOPEZ

DATE/TIME OF TEST: 01/16/2016 10:06:18

Diagnostics Check		OK	10:03	
Air Blank		0.000	10:03	
Control Test		0.079	10:04	
Air Blank		0.000	10:04	
Subject Sample	# 1		10:04	Subject Test Refused
Breath Volume		0.000	10:04	
Air Blank		0.000	10:05	
Control Test		0.079	10:05	
Air Blank		0.000	10:06	
Diagnostics Check		OK	10:06	

Data Download Date/Time: 01/29/2016 12:59

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESCOBAR,ALEJANDRO,M
BREATH TEST OPERATOR AGENCY: Coral Gables PD
ARRESTING OFFICER: LUACES,TEREASA
ARRESTING OFFICER AGENCY: Coral Gables PD
LAST AGENCY INSPECTION: 12/24/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12014080A1
DRY GAS EXPIRY DATE: 06/01/2016
VIOLATION CODE: DUI
ARREST TIME: 10:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:41

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 01/18/2016 11:17:33

Diagnostics Check		OK	11:14	
Air Blank		0.000	11:15	
Control Test		0.080	11:15	
Air Blank		0.000	11:16	
Subject Sample	# 1		11:16	Subject Test Refused
Breath Volume		0.000	11:16	
Air Blank		0.000	11:16	
Control Test		0.080	11:16	
Air Blank		0.000	11:17	
Diagnostics Check		OK	11:17	

Data Download Date/Time: 01/29/2016 12:59

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CABRAL,EDUARDO,
BREATH TEST OPERATOR AGENCY: Coral Gables PD
ARRESTING OFFICER: CABRAL,EDUARDO
ARRESTING OFFICER AGENCY: Coral Gables PD
LAST AGENCY INSPECTION: 12/24/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12014080A1
DRY GAS EXPIRY DATE: 06/01/2016
VIOLATION CODE: DUI
ARREST TIME: 20:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:31

SUBJECT'S LAST NAME: BAKER

DATE/TIME OF TEST: 01/20/2016 20:52:32

Diagnostics Check	OK	20:51	
Air Blank	0.000	20:51	
Control Test	0.072	20:51	Control Outside Tolerance
Air Blank	0.000	20:52	

Data Download Date/Time: 01/29/2016 12:59

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CABRAL,EDUARDO,
BREATH TEST OPERATOR AGENCY: Coral Gables PD
ARRESTING OFFICER: GUERRA,GABRIEL
ARRESTING OFFICER AGENCY: Coral Gables PD
LAST AGENCY INSPECTION: 12/24/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12014080A1
DRY GAS EXPIRY DATE: 06/01/2016
VIOLATION CODE: DUI
ARREST TIME: 20:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:31

SUBJECT'S LAST NAME: BAKER

DATE/TIME OF TEST: 01/20/2016 21:01:57

Diagnostics Check		OK	20:55
Air Blank		0.000	20:56
Control Test		0.080	20:56
Air Blank		0.000	20:56
Subject Sample	# 1	0.260	20:57
Breath Volume		3.386	20:57
Air Blank		0.000	20:58
Air Blank		0.000	20:59
Subject Sample	# 2	0.254	21:00
Breath Volume		2.257	21:00
Air Blank		0.000	21:01
Control Test		0.079	21:01
Air Blank		0.000	21:01
Diagnostics Check		OK	21:01

Data Download Date/Time: 01/29/2016 12:59

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GORDILLO,SAMUEL,E
BREATH TEST OPERATOR AGENCY: Coral Gables PD
ARRESTING OFFICER: MEDINA,MANUEL
ARRESTING OFFICER AGENCY: Coral Gables PD
LAST AGENCY INSPECTION: 12/24/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12014080A1
DRY GAS EXPIRY DATE: 06/01/2016
VIOLATION CODE: DUI
ARREST TIME: 20:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:49

SUBJECT'S LAST NAME: JAVECH

DATE/TIME OF TEST: 01/24/2016 22:15:57

Diagnostics Check		OK	22:12	
Air Blank		0.000	22:13	
Control Test		0.080	22:13	
Air Blank		0.000	22:13	
Subject Sample	# 1		22:14	Subject Test Refused
Breath Volume		0.000	22:14	
Air Blank		0.000	22:15	
Control Test		0.080	22:15	
Air Blank		0.000	22:15	
Diagnostics Check		OK	22:15	

Data Download Date/Time: 01/29/2016 12:59

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAHA,KHALED,
BREATH TEST OPERATOR AGENCY: Coral Gables PD
ARRESTING OFFICER: TAHA,KHALED
ARRESTING OFFICER AGENCY: Coral Gables PD
LAST AGENCY INSPECTION: 12/24/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12014080A1
DRY GAS EXPIRY DATE: 06/01/2016
VIOLATION CODE: DUI
ARREST TIME: 23:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:14

SUBJECT'S LAST NAME: DETRES

DATE/TIME OF TEST: 01/28/2016 00:36:56

Diagnostics Check	OK	00:36	
Air Blank		00:36	RFI Detect
Air Blank		00:36	RFI Detect

Data Download Date/Time: 01/29/2016 12:59

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000875

INSTRUMENT REGISTERED TO: CALHOUN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEMPLE,DARRYL,D
BREATH TEST OPERATOR AGENCY: Calhoun County SO
ARRESTING OFFICER: MCCORVEY,BEAU
ARRESTING OFFICER AGENCY: Calhoun County SO
LAST AGENCY INSPECTION: 01/08/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07014080A2
DRY GAS EXPIRY DATE: 05/01/2016
VIOLATION CODE: DUI
ARREST TIME: 17:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:27

SUBJECT'S LAST NAME: BOTTS

DATE/TIME OF TEST: 01/13/2016 18:07:39

Diagnostics Check		OK	17:59	
Air Blank		0.000	18:00	
Control Test		0.078	18:00	
Air Blank		0.000	18:01	
Subject Sample	# 1	0.203	18:01	
Breath Volume		2.261	18:01	
Air Blank		0.000	18:02	
Air Blank		0.000	18:04	
Subject Sample	# 2		18:07	Sequence Aborted
Breath Volume		0.000	18:07	
Air Blank		0.000	18:07	

Data Download Date/Time: 02/18/2016 09:53

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000876

INSTRUMENT REGISTERED TO: HIALEAH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARROJAS,GABRIEL,L
BREATH TEST OPERATOR AGENCY: Hialeah PD
ARRESTING OFFICER: VALDEZ,JUNIOR
ARRESTING OFFICER AGENCY: Hialeah PD
LAST AGENCY INSPECTION: 12/10/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06415080A1
DRY GAS EXPIRY DATE: 04/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:56

SUBJECT'S LAST NAME: GOZALEZ

DATE/TIME OF TEST: 01/01/2016 22:27:07

Diagnostics Check		OK	22:24	
Air Blank		0.000	22:24	
Control Test		0.078	22:25	
Air Blank		0.000	22:25	
Subject Sample	# 1		22:25	Subject Test Refused
Breath Volume		0.000	22:25	
Air Blank		0.000	22:26	
Control Test		0.077	22:26	
Air Blank		0.000	22:27	
Diagnostics Check		OK	22:27	

Data Download Date/Time: 01/21/2016 11:08

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000876

INSTRUMENT REGISTERED TO: HIALEAH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARROJAS,GABRIEL,L
BREATH TEST OPERATOR AGENCY: Hialeah PD
ARRESTING OFFICER: GUTIERREZ,RALAEL
ARRESTING OFFICER AGENCY: Hialeah PD
LAST AGENCY INSPECTION: 12/10/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06415080A1
DRY GAS EXPIRY DATE: 04/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:21

SUBJECT'S LAST NAME: FLORIAN

DATE/TIME OF TEST: 01/12/2016 21:49:04

Diagnostics Check		OK	21:42
Air Blank		0.000	21:42
Control Test		0.077	21:43
Air Blank		0.000	21:43
Subject Sample	# 1	0.146	21:44
Breath Volume		2.847	21:44
Air Blank		0.000	21:45
Air Blank		0.000	21:46
Subject Sample	# 2	0.150	21:47
Breath Volume		1.597	21:47
Air Blank		0.000	21:48
Control Test		0.076	21:48
Air Blank		0.000	21:48
Diagnostics Check		OK	21:49

Data Download Date/Time: 01/21/2016 11:08

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000876

INSTRUMENT REGISTERED TO: HIALEAH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ACOSTA,MARCELINO,R
BREATH TEST OPERATOR AGENCY: Hialeah PD
ARRESTING OFFICER: MONTALVO,B
ARRESTING OFFICER AGENCY: Hialeah PD
LAST AGENCY INSPECTION: 12/10/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06415080A1
DRY GAS EXPIRY DATE: 04/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: HARDEN

DATE/TIME OF TEST: 01/17/2016 01:35:34

Diagnostics Check		OK	01:28
Air Blank		0.000	01:28
Control Test		0.077	01:28
Air Blank		0.000	01:29
Subject Sample	# 1	0.158	01:30
Breath Volume		2.488	01:30
Air Blank		0.000	01:31
Air Blank		0.000	01:33
Subject Sample	# 2	0.157	01:33
Breath Volume		3.156	01:33
Air Blank		0.000	01:34
Control Test		0.077	01:34
Air Blank		0.000	01:35
Diagnostics Check		OK	01:35

Data Download Date/Time: 01/21/2016 11:08

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000878

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REYES,GEORGE,
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: MURGUIDO,ALEX
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 01/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1673300
DRY GAS EXPIRY DATE: 04/28/2017
VIOLATION CODE: DUI
ARREST TIME: 16:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:46

SUBJECT'S LAST NAME: CORTES-VASQUEZ

DATE/TIME OF TEST: 01/28/2016 17:17:46

Diagnostics Check		OK	17:07	
Air Blank		0.000	17:08	
Control Test		0.080	17:08	
Air Blank		0.000	17:08	
Subject Sample	# 1	0.248	17:12	Volume Not Met
Breath Volume		0.847	17:12	
Air Blank		0.000	17:13	
Air Blank		0.000	17:14	
Subject Sample	# 2		17:16	Subject Test Refused
Breath Volume		0.234	17:16	
Air Blank		0.000	17:16	
Control Test		0.080	17:17	
Air Blank		0.000	17:17	
Diagnostics Check		OK	17:17	

Data Download Date/Time: 02/29/2016 07:13

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000878

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEPHENSON,JENNIFER,L
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: STEPHENSON,JENNIFER
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 01/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1673300
DRY GAS EXPIRY DATE: 04/28/2017
VIOLATION CODE: DUI
ARREST TIME: 04:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:07

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 01/30/2016 05:45:28

Diagnostics Check		OK	05:39
Air Blank		0.000	05:39
Control Test		0.081	05:39
Air Blank		0.000	05:40
Subject Sample	# 1	0.198	05:40
Breath Volume		1.757	05:40
Air Blank		0.000	05:41
Air Blank		0.000	05:43
Subject Sample	# 2	0.206	05:43
Breath Volume		1.761	05:43
Air Blank		0.000	05:44
Control Test		0.081	05:44
Air Blank		0.000	05:45
Diagnostics Check		OK	05:45

Data Download Date/Time: 02/29/2016 07:13

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WENDLING,JOSHUA,A
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: ESPINOSA,HARRY
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 01/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 21:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:43

SUBJECT'S LAST NAME: VERA BETANCOURT

DATE/TIME OF TEST: 01/07/2016 22:12:58

Diagnostics Check		OK	22:09	
Air Blank		0.000	22:10	
Control Test		0.079	22:10	
Air Blank		0.000	22:11	
Subject Sample	# 1		22:11	Subject Test Refused
Breath Volume		0.000	22:11	
Air Blank		0.000	22:12	
Control Test		0.079	22:12	
Air Blank		0.000	22:12	
Diagnostics Check		OK	22:12	

Data Download Date/Time: 02/12/2016 09:32

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WENDLING,JOSHUA,A
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: ESPINOSA,HARRY
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 01/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 21:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:43

SUBJECT'S LAST NAME: VERA BETANCOURT

DATE/TIME OF TEST: 01/07/2016 22:29:33

Diagnostics Check		OK	22:23
Air Blank		0.000	22:23
Control Test		0.080	22:23
Air Blank		0.000	22:24
Subject Sample	# 1	0.165	22:25
Breath Volume		1.175	22:25
Air Blank		0.000	22:25
Air Blank		0.000	22:27
Subject Sample	# 2	0.153	22:28
Breath Volume		1.335	22:28
Air Blank		0.000	22:28
Control Test		0.079	22:28
Air Blank		0.000	22:29
Diagnostics Check		OK	22:29

Data Download Date/Time: 02/12/2016 09:32

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURNS,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Aventura PD
ARRESTING OFFICER: BURNS,JEFFREY
ARRESTING OFFICER AGENCY: Aventura PD
LAST AGENCY INSPECTION: 01/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 07:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:50

SUBJECT'S LAST NAME: ZABALA

DATE/TIME OF TEST: 01/11/2016 08:17:38

Diagnostics Check		OK	08:11
Air Blank		0.000	08:12
Control Test		0.080	08:12
Air Blank		0.000	08:12
Subject Sample	# 1	0.000	08:13
Breath Volume		1.875	08:13
Air Blank		0.000	08:13
Air Blank		0.000	08:15
Subject Sample	# 2	0.000	08:16
Breath Volume		1.726	08:16
Air Blank		0.000	08:16
Control Test		0.079	08:17
Air Blank		0.000	08:17
Diagnostics Check		OK	08:17

Data Download Date/Time: 02/12/2016 09:32

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRAVIS,CHARLES,B
BREATH TEST OPERATOR AGENCY: Aventura PD
ARRESTING OFFICER: TRAVIS,CHARLES
ARRESTING OFFICER AGENCY: Aventura PD
LAST AGENCY INSPECTION: 01/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 10:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:28

SUBJECT'S LAST NAME: BELLO

DATE/TIME OF TEST: 01/19/2016 10:56:28

Diagnostics Check		OK	10:49	
Air Blank		0.000	10:50	
Control Test		0.080	10:50	
Air Blank		0.000	10:51	
Subject Sample	# 1	0.000	10:52	
Breath Volume		1.289	10:52	
Air Blank		0.000	10:52	
Air Blank		0.000	10:54	
Subject Sample	# 2		10:55	RFI Detect
Breath Volume		1.203	10:55	
Air Blank		0.000	10:55	
Control Test		0.079	10:55	
Air Blank		0.000	10:56	
Diagnostics Check		OK	10:56	

Data Download Date/Time: 02/12/2016 09:32

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRAVIS,CHARLES,B
BREATH TEST OPERATOR AGENCY: Aventura PD
ARRESTING OFFICER: TRAVIS,CHARLES
ARRESTING OFFICER AGENCY: Aventura PD
LAST AGENCY INSPECTION: 01/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 10:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:28

SUBJECT'S LAST NAME: BELLO

DATE/TIME OF TEST: 01/19/2016 11:02:30

Diagnostics Check		OK	10:58	
Air Blank		0.000	10:59	
Control Test		0.080	10:59	
Air Blank		0.000	10:59	
Subject Sample	# 1		11:01	Subject Test Refused
Breath Volume		0.039	11:01	
Air Blank		0.000	11:01	
Control Test		0.081	11:01	
Air Blank		0.000	11:02	
Diagnostics Check		OK	11:02	

Data Download Date/Time: 02/12/2016 09:32

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AGINS,LYLE,P
BREATH TEST OPERATOR AGENCY: Sunny Isles Beach PD
ARRESTING OFFICER: AGINS,LYLE
ARRESTING OFFICER AGENCY: Sunny Isles Beach PD
LAST AGENCY INSPECTION: 01/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:12

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 01/20/2016 03:38:50

Diagnostics Check		OK	03:32
Air Blank		0.000	03:32
Control Test		0.080	03:33
Air Blank		0.000	03:33
Subject Sample	# 1	0.179	03:34
Breath Volume		3.039	03:34
Air Blank		0.000	03:35
Air Blank		0.000	03:36
Subject Sample	# 2	0.182	03:37
Breath Volume		2.589	03:37
Air Blank		0.000	03:37
Control Test		0.079	03:38
Air Blank		0.000	03:38
Diagnostics Check		OK	03:38

Data Download Date/Time: 02/12/2016 09:32

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AGINS,LYLE,P
BREATH TEST OPERATOR AGENCY: Sunny Isles Beach PD
ARRESTING OFFICER: AGINS,LYLE
ARRESTING OFFICER AGENCY: Sunny Isles Beach PD
LAST AGENCY INSPECTION: 01/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:08

SUBJECT'S LAST NAME: KOSTYUCHENKO

DATE/TIME OF TEST: 01/24/2016 06:32:08

Diagnostics Check		OK	06:29	
Air Blank		0.000	06:29	
Control Test		0.081	06:30	
Air Blank		0.000	06:30	
Subject Sample	# 1		06:30	Subject Test Refused
Breath Volume		0.000	06:30	
Air Blank		0.000	06:31	
Control Test		0.081	06:31	
Air Blank		0.000	06:32	
Diagnostics Check		OK	06:32	

Data Download Date/Time: 02/12/2016 09:32

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHAPIRO,SABINA,M
BREATH TEST OPERATOR AGENCY: Aventura PD
ARRESTING OFFICER: SHAPIRO,SABINA
ARRESTING OFFICER AGENCY: Aventura PD
LAST AGENCY INSPECTION: 01/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 18:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:15

SUBJECT'S LAST NAME: GUY

DATE/TIME OF TEST: 01/30/2016 19:38:41

Diagnostics Check		OK	19:35	
Air Blank		0.000	19:36	
Control Test		0.079	19:36	
Air Blank		0.000	19:37	
Subject Sample	# 1		19:37	Subject Test Refused
Breath Volume		0.000	19:37	
Air Blank		0.000	19:37	
Control Test		0.080	19:38	
Air Blank		0.000	19:38	
Diagnostics Check		OK	19:38	

Data Download Date/Time: 02/12/2016 09:32

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCGRATH,ROBERT,E
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: TIPPENHAUER,RICHARD
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 01/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 08:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:25

SUBJECT'S LAST NAME: TORRES GONZALEZ

DATE/TIME OF TEST: 01/31/2016 09:52:42

Diagnostics Check		OK	09:45
Air Blank		0.000	09:46
Control Test		0.080	09:46
Air Blank		0.000	09:47
Subject Sample	# 1	0.200	09:48
Breath Volume		2.460	09:48
Air Blank		0.000	09:48
Air Blank		0.000	09:50
Subject Sample	# 2	0.194	09:51
Breath Volume		2.230	09:51
Air Blank		0.000	09:51
Control Test		0.080	09:52
Air Blank		0.000	09:52
Diagnostics Check		OK	09:52

Data Download Date/Time: 02/12/2016 09:32

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000881

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAITE,JASON,B
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: WAITE,JASON
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 01/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 15:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:52

SUBJECT'S LAST NAME: VENEGAS

DATE/TIME OF TEST: 01/09/2016 16:15:30

Diagnostics Check		OK	16:12	
Air Blank		0.000	16:12	
Control Test		0.080	16:13	
Air Blank		0.000	16:13	
Subject Sample	# 1		16:14	Subject Test Refused
Breath Volume		0.000	16:14	
Air Blank		0.000	16:14	
Control Test		0.080	16:14	
Air Blank		0.000	16:15	
Diagnostics Check		OK	16:15	

Data Download Date/Time: 02/05/2016 09:33

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000881

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREGORY,BRIAN,K
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: GREGORY,BRIAN
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 01/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 21:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:08

SUBJECT'S LAST NAME: LUIS

DATE/TIME OF TEST: 01/22/2016 22:32:08

Diagnostics Check		OK	22:29	
Air Blank		0.000	22:29	
Control Test		0.081	22:29	
Air Blank		0.000	22:30	
Subject Sample	# 1		22:30	Subject Test Refused
Breath Volume		0.000	22:30	
Air Blank		0.000	22:31	
Control Test		0.080	22:31	
Air Blank		0.000	22:32	
Diagnostics Check		OK	22:32	

Data Download Date/Time: 02/05/2016 09:33

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000881

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,STEVEN,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: ADAMS,STEVEN
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 01/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 22:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: LAVERDE

DATE/TIME OF TEST: 01/29/2016 23:52:16

Diagnostics Check		OK	23:45
Air Blank		0.000	23:46
Control Test		0.080	23:46
Air Blank		0.000	23:46
Subject Sample	# 1	0.097	23:47
Breath Volume		5.726	23:47
Air Blank		0.000	23:48
Air Blank		0.000	23:50
Subject Sample	# 2	0.090	23:50
Breath Volume		5.097	23:50
Air Blank		0.000	23:51
Control Test		0.081	23:51
Air Blank		0.000	23:52
Diagnostics Check		OK	23:52

Data Download Date/Time: 02/05/2016 09:33

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000881

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,STEVEN,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: ADAMS,STEVEN
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 01/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 01/31/2016 03:33:46

Diagnostics Check		OK	03:26
Air Blank		0.000	03:26
Control Test		0.081	03:26
Air Blank		0.000	03:27
Subject Sample	# 1	0.083	03:28
Breath Volume		1.417	03:28
Air Blank		0.000	03:28
Air Blank		0.000	03:30
Subject Sample	# 2	0.081	03:32
Breath Volume		1.246	03:32
Air Blank		0.000	03:32
Control Test		0.081	03:33
Air Blank		0.000	03:33
Diagnostics Check		OK	03:33

Data Download Date/Time: 02/05/2016 09:33

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEY,LAZARO,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: SANCHEZ,CHRISTOPHER
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 01/20/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: MENDEZ-SOLORSANO

DATE/TIME OF TEST: 01/30/2016 00:00:32

Diagnostics Check		OK	23:44	
Air Blank		0.000	23:45	
Control Test		0.078	23:45	
Air Blank		0.000	23:45	
Subject Sample	# 1	0.230	23:49	Volume Not Met
Breath Volume		0.394	23:49	
Air Blank		0.000	23:49	
Air Blank		0.000	23:51	
Subject Sample	# 2	0.249	23:54	No .020 agreement
Breath Volume		1.421	23:54	
Air Blank		0.000	23:54	
Air Blank		0.000	23:56	
Subject Sample	# 3	0.220	23:58	No .020 agreement
Breath Volume		1.148	23:58	
Air Blank		0.000	23:59	
Control Test		0.076	23:59	
Air Blank		0.000	00:00	
Diagnostics Check		OK	00:00	

Data Download Date/Time: 02/05/2016 09:36

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEY,LAZARO,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 01/20/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 00:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:54

SUBJECT'S LAST NAME: CASTILLO

DATE/TIME OF TEST: 01/30/2016 01:28:00

Diagnostics Check		OK	01:25	
Air Blank		0.000	01:25	
Control Test		0.075	01:25	
Air Blank		0.000	01:26	
Subject Sample	# 1		01:26	Subject Test Refused
Breath Volume		0.000	01:26	
Air Blank		0.000	01:27	
Control Test		0.077	01:27	
Air Blank		0.000	01:27	
Diagnostics Check		OK	01:28	

Data Download Date/Time: 02/05/2016 09:36

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEY,LAZARO,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: FERRO,LISA
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 01/20/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:31

SUBJECT'S LAST NAME: LAGO BRINONES

DATE/TIME OF TEST: 01/30/2016 03:18:36

Diagnostics Check		OK	03:08	
Air Blank		0.000	03:09	
Control Test		0.078	03:09	
Air Blank		0.000	03:10	
Subject Sample	# 1	0.096	03:11	
Breath Volume		1.496	03:11	
Air Blank		0.000	03:12	
Air Blank		0.000	03:14	
Subject Sample	# 2		03:17	Subject Test Refused
Breath Volume		0.539	03:17	
Air Blank		0.000	03:17	
Control Test		0.078	03:18	
Air Blank		0.000	03:18	
Diagnostics Check		OK	03:18	

Data Download Date/Time: 02/05/2016 09:36

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEY,LAZARO,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: FERRO,LISA
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 01/20/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:31

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 01/31/2016 03:26:41

Diagnostics Check		OK	03:23	
Air Blank		0.000	03:24	
Control Test		0.077	03:24	
Air Blank		0.000	03:24	
Subject Sample	# 1		03:25	Subject Test Refused
Breath Volume		0.000	03:25	
Air Blank		0.000	03:25	
Control Test		0.077	03:26	
Air Blank		0.000	03:26	
Diagnostics Check		OK	03:26	

Data Download Date/Time: 02/05/2016 09:36

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESPINOZA,PABLO,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: FRANCO,GREGORY
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 01/20/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 04:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:15

SUBJECT'S LAST NAME: DE GRACIA

DATE/TIME OF TEST: 01/31/2016 04:43:15

Diagnostics Check		OK	04:36
Air Blank		0.000	04:37
Control Test		0.078	04:37
Air Blank		0.000	04:37
Subject Sample	# 1	0.101	04:38
Breath Volume		3.449	04:38
Air Blank		0.000	04:39
Air Blank		0.000	04:40
Subject Sample	# 2	0.105	04:41
Breath Volume		2.933	04:41
Air Blank		0.000	04:42
Control Test		0.078	04:42
Air Blank		0.000	04:43
Diagnostics Check		OK	04:43

Data Download Date/Time: 02/05/2016 09:36

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEL CASTILLO,RAUL,A
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: NOVO,ERICK
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 12:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:19

SUBJECT'S LAST NAME: GUTIERREZ

DATE/TIME OF TEST: 01/01/2016 13:42:46

Diagnostics Check		OK	13:39	
Air Blank		0.000	13:40	
Control Test		0.079	13:40	
Air Blank		0.000	13:40	
Subject Sample	# 1		13:41	Subject Test Refused
Breath Volume		0.000	13:41	
Air Blank		0.000	13:41	
Control Test		0.078	13:42	
Air Blank		0.000	13:42	
Diagnostics Check		OK	13:42	

Data Download Date/Time: 01/21/2016 09:37

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VARGAS,D,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: VARGAS,D
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 02:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:46

SUBJECT'S LAST NAME: QUINTANA

DATE/TIME OF TEST: 01/02/2016 03:14:45

Diagnostics Check		OK	03:08
Air Blank		0.000	03:08
Control Test		0.080	03:09
Air Blank		0.000	03:09
Subject Sample	# 1	0.171	03:10
Breath Volume		2.511	03:10
Air Blank		0.000	03:10
Air Blank		0.000	03:12
Subject Sample	# 2	0.173	03:13
Breath Volume		2.125	03:13
Air Blank		0.000	03:13
Control Test		0.078	03:14
Air Blank		0.000	03:14
Diagnostics Check		OK	03:14

Data Download Date/Time: 01/21/2016 09:37

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANDREOZZI,ADAM,R
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: ANDREOZZI,ADAM
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 01:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: IZARZUGAZA

DATE/TIME OF TEST: 01/06/2016 02:39:55

Diagnostics Check		OK	02:32
Air Blank		0.000	02:33
Control Test		0.079	02:33
Air Blank		0.000	02:34
Subject Sample	# 1	0.183	02:35
Breath Volume		2.796	02:35
Air Blank		0.000	02:35
Air Blank		0.000	02:37
Subject Sample	# 2	0.177	02:38
Breath Volume		2.324	02:38
Air Blank		0.000	02:38
Control Test		0.077	02:39
Air Blank		0.000	02:39
Diagnostics Check		OK	02:39

Data Download Date/Time: 01/21/2016 09:37

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,CHRISTINA,A
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: MARTIN,CHRISTINA
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: ORDONEZ JR

DATE/TIME OF TEST: 01/07/2016 00:37:11

Diagnostics Check		OK	00:30
Air Blank		0.000	00:31
Control Test		0.079	00:31
Air Blank		0.000	00:32
Subject Sample	# 1	0.000	00:32
Breath Volume		3.113	00:32
Air Blank		0.000	00:33
Air Blank		0.000	00:35
Subject Sample	# 2	0.000	00:35
Breath Volume		4.058	00:35
Air Blank		0.000	00:36
Control Test		0.077	00:36
Air Blank		0.000	00:37
Diagnostics Check		OK	00:37

Data Download Date/Time: 01/21/2016 09:37

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARDELAS,MAYLEE,
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: BARDELAS,MAYLEE
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:54

SUBJECT'S LAST NAME: REYES-MORENO

DATE/TIME OF TEST: 01/09/2016 03:23:45

Diagnostics Check		OK	03:17
Air Blank		0.000	03:17
Control Test		0.079	03:18
Air Blank		0.000	03:18
Subject Sample	# 1	0.090	03:19
Breath Volume		3.183	03:19
Air Blank		0.000	03:19
Air Blank		0.000	03:21
Subject Sample	# 2	0.086	03:22
Breath Volume		2.222	03:22
Air Blank		0.000	03:22
Control Test		0.077	03:23
Air Blank		0.000	03:23
Diagnostics Check		OK	03:23

Data Download Date/Time: 01/21/2016 09:37

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARDELAS,MAYLE,
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: DESIUS,DUCLAIRE
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: MUNOZ-RODRIGUEZ

DATE/TIME OF TEST: 01/09/2016 03:54:46

Diagnostics Check		OK	03:47
Air Blank		0.000	03:48
Control Test		0.079	03:48
Air Blank		0.000	03:48
Subject Sample	# 1	0.149	03:50
Breath Volume		2.523	03:50
Air Blank		0.000	03:50
Air Blank		0.000	03:52
Subject Sample	# 2	0.140	03:53
Breath Volume		1.570	03:53
Air Blank		0.000	03:53
Control Test		0.075	03:54
Air Blank		0.000	03:54
Diagnostics Check		OK	03:54

Data Download Date/Time: 01/21/2016 09:37

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VARGAS,D,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: VARGAS,D
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 03:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:46

SUBJECT'S LAST NAME: PEREZ GONZALEZ

DATE/TIME OF TEST: 01/09/2016 04:19:46

Diagnostics Check		OK	04:12
Air Blank		0.000	04:13
Control Test		0.078	04:13
Air Blank		0.000	04:14
Subject Sample	# 1	0.164	04:14
Breath Volume		2.242	04:14
Air Blank		0.000	04:15
Air Blank		0.000	04:17
Subject Sample	# 2	0.179	04:18
Breath Volume		2.296	04:18
Air Blank		0.000	04:18
Control Test		0.078	04:19
Air Blank		0.000	04:19
Diagnostics Check		OK	04:19

Data Download Date/Time: 01/21/2016 09:37

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: VENTO LUIS

DATE/TIME OF TEST: 01/09/2016 04:41:08

Diagnostics Check		OK	04:38	
Air Blank		0.000	04:38	
Control Test		0.078	04:38	
Air Blank		0.000	04:39	
Subject Sample	# 1		04:39	Subject Test Refused
Breath Volume		0.000	04:39	
Air Blank		0.000	04:40	
Control Test		0.078	04:40	
Air Blank		0.000	04:41	
Diagnostics Check		OK	04:41	

Data Download Date/Time: 01/21/2016 09:37

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VALENCIA,KEVIN,A
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: VALLADARES,CARLOS
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 21:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:55

SUBJECT'S LAST NAME: GARIBOLDI

DATE/TIME OF TEST: 01/10/2016 22:25:34

Diagnostics Check		OK	22:15	
Air Blank		0.000	22:16	
Control Test		0.080	22:16	
Air Blank		0.000	22:17	
Subject Sample	# 1	0.229	22:20	Slope Not Level
Breath Volume		1.609	22:20	
Air Blank		0.000	22:21	
Air Blank		0.000	22:22	
Subject Sample	# 2	0.243	22:24	
Breath Volume		2.878	22:24	
Air Blank			22:25	Purge Fail
Air Blank		0.000	22:25	

Data Download Date/Time: 01/21/2016 09:37

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VALENCIA,KEVIN,A
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: VALLADARES,CARLOS
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 21:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:55

SUBJECT'S LAST NAME: GARIBOLDI

DATE/TIME OF TEST: 01/10/2016 22:37:51

Diagnostics Check		OK	22:28	
Air Blank		0.000	22:29	
Control Test		0.077	22:29	
Air Blank		0.000	22:29	
Subject Sample	# 1	0.246	22:31	
Breath Volume		2.761	22:31	
Air Blank		0.000	22:32	
Air Blank		0.000	22:33	
Subject Sample	# 2	0.240	22:36	
Breath Volume		2.753	22:36	
Air Blank			22:37	Purge Fail
Air Blank		0.000	22:37	

Data Download Date/Time: 01/21/2016 09:37

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESCONTRELA,CRISTINA,
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: BARDELAS,MAYLE
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 01:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: RANGEL TAMAYO

DATE/TIME OF TEST: 01/11/2016 01:51:58

Diagnostics Check	OK	01:50	
Air Blank	0.000	01:51	
Control Test	0.078	01:51	
Air Blank		01:51	RFI Detect
Air Blank		01:51	RFI Detect

Data Download Date/Time: 01/21/2016 09:37

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESCONTRELA,CRISTINA,
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: BARDELAS,MAYLE
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 01:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: RANGEL TAMAYO

DATE/TIME OF TEST: 01/11/2016 01:58:38

Diagnostics Check		OK	01:55	
Air Blank		0.000	01:56	
Control Test		0.078	01:56	
Air Blank		0.000	01:57	
Subject Sample	# 1		01:57	Subject Test Refused
Breath Volume		0.000	01:57	
Air Blank		0.000	01:57	
Control Test		0.078	01:58	
Air Blank		0.000	01:58	
Diagnostics Check		OK	01:58	

Data Download Date/Time: 01/21/2016 09:37

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VARGAS,D,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: VARGAS,D
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:18

SUBJECT'S LAST NAME: ROJAS

DATE/TIME OF TEST: 01/14/2016 02:51:13

Diagnostics Check		OK	02:43
Air Blank		0.000	02:43
Control Test		0.079	02:44
Air Blank		0.000	02:44
Subject Sample	# 1	0.121	02:46
Breath Volume		1.343	02:46
Air Blank		0.000	02:46
Air Blank		0.000	02:48
Subject Sample	# 2	0.126	02:49
Breath Volume		1.550	02:49
Air Blank		0.000	02:50
Control Test		0.078	02:50
Air Blank		0.000	02:51
Diagnostics Check		OK	02:51

Data Download Date/Time: 01/21/2016 09:37

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VALENCIA,KEVIN,A
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: VALENCIA,KEVIN
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 22:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:07

SUBJECT'S LAST NAME: LINDSAY

DATE/TIME OF TEST: 01/14/2016 23:37:44

Diagnostics Check		OK	23:27	
Air Blank		0.000	23:27	
Control Test		0.079	23:28	
Air Blank		0.000	23:28	
Subject Sample	# 1		23:31	No Sample Provided
Breath Volume		0.000	23:31	
Air Blank		0.000	23:32	
Air Blank		0.000	23:34	
Subject Sample	# 2		23:36	Subject Test Refused
Breath Volume		0.000	23:36	
Air Blank		0.000	23:36	
Control Test		0.079	23:37	
Air Blank		0.000	23:37	
Diagnostics Check		OK	23:37	

Data Download Date/Time: 01/21/2016 09:37

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: VARGAS,DAEL
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 03:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:28

SUBJECT'S LAST NAME: VENTURA

DATE/TIME OF TEST: 01/15/2016 04:56:10

Diagnostics Check		OK	04:49
Air Blank		0.000	04:50
Control Test		0.080	04:50
Air Blank		0.000	04:50
Subject Sample	# 1	0.106	04:51
Breath Volume		1.917	04:51
Air Blank		0.000	04:52
Air Blank		0.000	04:54
Subject Sample	# 2	0.102	04:54
Breath Volume		1.394	04:54
Air Blank		0.000	04:55
Control Test		0.077	04:55
Air Blank		0.000	04:56
Diagnostics Check		OK	04:56

Data Download Date/Time: 01/21/2016 09:37

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,EDWIN,M
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: PEREZ,EDWIN
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 06:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:17

SUBJECT'S LAST NAME: SIERRA

DATE/TIME OF TEST: 01/15/2016 07:41:04

Diagnostics Check		OK	07:38	
Air Blank		0.000	07:38	
Control Test		0.080	07:39	
Air Blank		0.000	07:39	
Subject Sample	# 1		07:39	Subject Test Refused
Breath Volume		0.000	07:39	
Air Blank		0.000	07:40	
Control Test		0.079	07:40	
Air Blank		0.000	07:40	
Diagnostics Check		OK	07:41	

Data Download Date/Time: 01/21/2016 09:37

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VARGAS,D,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: VARGAS,D
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 02:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:14

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 01/16/2016 03:45:24

Diagnostics Check		OK	03:38
Air Blank		0.000	03:39
Control Test		0.080	03:39
Air Blank		0.000	03:40
Subject Sample	# 1	0.209	03:40
Breath Volume		4.269	03:40
Air Blank		0.000	03:41
Air Blank		0.000	03:43
Subject Sample	# 2	0.210	03:43
Breath Volume		3.207	03:43
Air Blank		0.000	03:44
Control Test		0.078	03:44
Air Blank		0.000	03:45
Diagnostics Check		OK	03:45

Data Download Date/Time: 01/21/2016 09:37

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: MYRTIL,R
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 04:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:25

SUBJECT'S LAST NAME: CORDERO

DATE/TIME OF TEST: 01/16/2016 06:37:21

Diagnostics Check		OK	06:23	
Air Blank		0.000	06:23	
Control Test		0.079	06:24	
Air Blank		0.000	06:24	
Subject Sample	# 1	0.166	06:27	Volume Not Met
Breath Volume		0.820	06:27	
Air Blank		0.000	06:28	
Air Blank		0.000	06:30	
Subject Sample	# 2	0.162	06:31	
Breath Volume		1.476	06:31	
Air Blank		0.000	06:32	
Air Blank		0.000	06:34	
Subject Sample	# 3	0.165	06:35	
Breath Volume		1.691	06:35	
Air Blank		0.000	06:36	
Control Test		0.078	06:36	
Air Blank		0.000	06:37	
Diagnostics Check		OK	06:37	

Data Download Date/Time: 01/21/2016 09:37

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BONILLA,JOSE,I
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: BONILLA,JOSE
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 06:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:23

SUBJECT'S LAST NAME: CALAFELL JR

DATE/TIME OF TEST: 01/16/2016 06:47:13

Diagnostics Check		OK	06:44	
Air Blank		0.000	06:44	
Control Test		0.078	06:45	
Air Blank		0.000	06:45	
Subject Sample	# 1		06:45	Subject Test Refused
Breath Volume		0.000	06:45	
Air Blank		0.000	06:46	
Control Test		0.079	06:46	
Air Blank		0.000	06:47	
Diagnostics Check		OK	06:47	

Data Download Date/Time: 01/21/2016 09:37

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYRTIL,REGINALD,J
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: KINSEYSMITH,FLELICIA
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:54

SUBJECT'S LAST NAME: BAKER

DATE/TIME OF TEST: 01/21/2016 02:35:54

Diagnostics Check		OK	02:28
Air Blank		0.000	02:28
Control Test		0.080	02:28
Air Blank		0.000	02:29
Subject Sample	# 1	0.176	02:31
Breath Volume		1.164	02:31
Air Blank		0.000	02:31
Air Blank		0.000	02:33
Subject Sample	# 2	0.175	02:34
Breath Volume		1.101	02:34
Air Blank		0.000	02:34
Control Test		0.077	02:35
Air Blank		0.000	02:35
Diagnostics Check		OK	02:35

Data Download Date/Time: 01/21/2016 09:37

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYRTIL,REGINALD,J
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: VARGAS,DAEL
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 03:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:48

SUBJECT'S LAST NAME: CAMPBELL

DATE/TIME OF TEST: 01/21/2016 04:13:40

Diagnostics Check		OK	04:10	
Air Blank		0.000	04:11	
Control Test		0.080	04:11	
Air Blank		0.000	04:12	
Subject Sample	# 1		04:12	Subject Test Refused
Breath Volume		0.000	04:12	
Air Blank		0.000	04:12	
Control Test		0.079	04:13	
Air Blank		0.000	04:13	
Diagnostics Check		OK	04:13	

Data Download Date/Time: 01/21/2016 09:37

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYRTIL,REGINALD,J
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 01/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 01:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:34

SUBJECT'S LAST NAME: SANDS

DATE/TIME OF TEST: 01/22/2016 02:30:36

Diagnostics Check		OK	02:21	
Air Blank		0.000	02:22	
Control Test		0.079	02:22	
Air Blank		0.000	02:23	
Subject Sample	# 1	0.227	02:25	
Breath Volume		2.039	02:25	
Air Blank		0.000	02:26	
Air Blank		0.000	02:28	
Subject Sample	# 2	0.231	02:28	
Breath Volume		1.253	02:28	
Air Blank		0.000	02:29	
Control Test		0.074	02:29	Control Outside Tolerance
Air Blank		0.000	02:30	

Data Download Date/Time: 02/04/2016 11:23

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYRTIL,REGINALD,J
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 01/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 01:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:34

SUBJECT'S LAST NAME: SANDS

DATE/TIME OF TEST: 01/22/2016 02:42:27

Diagnostics Check		OK	02:34
Air Blank		0.000	02:35
Control Test		0.076	02:35
Air Blank		0.000	02:35
Subject Sample	# 1	0.227	02:37
Breath Volume		1.425	02:37
Air Blank		0.000	02:38
Air Blank		0.000	02:39
Subject Sample	# 2	0.223	02:40
Breath Volume		1.875	02:40
Air Blank		0.000	02:41
Control Test		0.076	02:41
Air Blank		0.000	02:42
Diagnostics Check		OK	02:42

Data Download Date/Time: 02/04/2016 11:23

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 01/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 05:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:35

SUBJECT'S LAST NAME: CARDONA

DATE/TIME OF TEST: 01/22/2016 06:20:00

Diagnostics Check		OK	06:17	
Air Blank		0.000	06:17	
Control Test		0.078	06:17	
Air Blank		0.000	06:18	
Subject Sample	# 1		06:18	Subject Test Refused
Breath Volume		0.000	06:18	
Air Blank		0.000	06:19	
Control Test		0.077	06:19	
Air Blank		0.000	06:19	
Diagnostics Check		OK	06:20	

Data Download Date/Time: 02/04/2016 11:23

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYRTIL,REGINALD,J
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: FRANCO,GREGORY
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 01/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 03:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:39

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 01/23/2016 04:09:21

Diagnostics Check		OK	04:02
Air Blank		0.000	04:03
Control Test		0.079	04:03
Air Blank		0.000	04:03
Subject Sample	# 1	0.145	04:04
Breath Volume		2.480	04:04
Air Blank		0.000	04:05
Air Blank		0.000	04:06
Subject Sample	# 2	0.145	04:07
Breath Volume		3.328	04:07
Air Blank		0.000	04:08
Control Test		0.078	04:08
Air Blank		0.000	04:09
Diagnostics Check		OK	04:09

Data Download Date/Time: 02/04/2016 11:23

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRANCO,GREGORY,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 01/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 03:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 01/23/2016 04:38:53

Diagnostics Check		OK	04:32
Air Blank		0.000	04:33
Control Test		0.079	04:33
Air Blank		0.000	04:33
Subject Sample	# 1	0.177	04:34
Breath Volume		3.414	04:34
Air Blank		0.000	04:35
Air Blank		0.000	04:36
Subject Sample	# 2	0.184	04:37
Breath Volume		3.589	04:37
Air Blank		0.000	04:37
Control Test		0.077	04:38
Air Blank		0.000	04:38
Diagnostics Check		OK	04:38

Data Download Date/Time: 02/04/2016 11:23

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 01/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 05:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:30

SUBJECT'S LAST NAME: MEDAL

DATE/TIME OF TEST: 01/23/2016 07:06:04

Diagnostics Check		OK	06:59
Air Blank		0.000	07:00
Control Test		0.078	07:00
Air Blank		0.000	07:00
Subject Sample	# 1	0.146	07:01
Breath Volume		1.589	07:01
Air Blank		0.000	07:02
Air Blank		0.000	07:03
Subject Sample	# 2	0.161	07:04
Breath Volume		1.792	07:04
Air Blank		0.000	07:05
Control Test		0.077	07:05
Air Blank		0.000	07:05
Diagnostics Check		OK	07:06

Data Download Date/Time: 02/04/2016 11:24

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANDREOZZI,ADAM,R
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: VALES,JEAN
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 01/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 21:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: GARAY

DATE/TIME OF TEST: 01/23/2016 22:39:05

Diagnostics Check		OK	22:35	
Air Blank		0.000	22:36	
Control Test		0.078	22:36	
Air Blank		0.000	22:37	
Subject Sample	# 1		22:37	Subject Test Refused
Breath Volume		0.000	22:37	
Air Blank		0.000	22:38	
Control Test		0.079	22:38	
Air Blank		0.000	22:39	
Diagnostics Check		OK	22:39	

Data Download Date/Time: 02/04/2016 11:24

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIANCHI,MARIO,D
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: BIANCHI,MARIO
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 01/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 22:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:52

SUBJECT'S LAST NAME: HERNANDEZ MARAGOTO

DATE/TIME OF TEST: 01/24/2016 23:43:48

Diagnostics Check		OK	23:30	
Air Blank		0.000	23:30	
Control Test		0.080	23:30	
Air Blank		0.000	23:31	
Subject Sample	# 1	0.209	23:34	Volume Not Met
Breath Volume		0.183	23:34	
Air Blank		0.000	23:35	
Air Blank		0.000	23:37	
Subject Sample	# 2	0.201	23:38	No .020 agreement
Breath Volume		1.871	23:38	
Air Blank		0.000	23:39	
Air Blank		0.000	23:41	
Subject Sample	# 3	0.222	23:41	No .020 agreement
Breath Volume		1.312	23:41	
Air Blank		0.000	23:42	
Control Test		0.075	23:43	
Air Blank		0.000	23:43	
Diagnostics Check		OK	23:43	

Data Download Date/Time: 02/04/2016 11:24

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIANCHI,MARIO,D
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: BIANCHI,MARIO
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 01/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 22:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:52

SUBJECT'S LAST NAME: HERNANDEZ MARAGOTO

DATE/TIME OF TEST: 01/25/2016 00:28:58

Diagnostics Check		OK	00:21
Air Blank		0.000	00:22
Control Test		0.078	00:22
Air Blank		0.000	00:23
Subject Sample	# 1	0.217	00:24
Breath Volume		1.839	00:24
Air Blank		0.000	00:24
Air Blank		0.000	00:26
Subject Sample	# 2	0.217	00:27
Breath Volume		2.890	00:27
Air Blank		0.000	00:28
Control Test		0.078	00:28
Air Blank		0.000	00:28
Diagnostics Check		OK	00:28

Data Download Date/Time: 02/04/2016 11:24

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUFFNER,HAL,E
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: RUFFNER,HAL
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 01/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 03:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:31

SUBJECT'S LAST NAME: PEREZ-CORONA

DATE/TIME OF TEST: 01/25/2016 05:09:39

Diagnostics Check		OK	05:03
Air Blank		0.000	05:03
Control Test		0.079	05:03
Air Blank		0.000	05:04
Subject Sample	# 1	0.188	05:04
Breath Volume		3.578	05:04
Air Blank		0.000	05:05
Air Blank		0.000	05:07
Subject Sample	# 2	0.179	05:07
Breath Volume		3.437	05:07
Air Blank		0.000	05:08
Control Test		0.077	05:09
Air Blank		0.000	05:09
Diagnostics Check		OK	05:09

Data Download Date/Time: 02/04/2016 11:24

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TELLEZ,ALBERTO,
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: TELLEZ,ALBERTO
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 01/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 20:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:10

SUBJECT'S LAST NAME: D630292738700

DATE/TIME OF TEST: 01/25/2016 21:33:00

Diagnostics Check	OK	21:32	
Air Blank		21:32	Ambient Fail
Air Blank	0.000	21:33	

Data Download Date/Time: 02/04/2016 11:24

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TELLEZ,ALBERTO,
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: TELLEZ,ALBERTO
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 01/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 20:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:10

SUBJECT'S LAST NAME: D630292738700

DATE/TIME OF TEST: 01/25/2016 21:40:25

Diagnostics Check		OK	21:35	
Air Blank		0.000	21:36	
Control Test		0.078	21:36	
Air Blank		0.000	21:36	
Subject Sample	# 1	0.249	21:39	
Breath Volume		1.359	21:39	
Air Blank			21:39	Purge Fail
Air Blank		0.000	21:40	

Data Download Date/Time: 02/04/2016 11:24

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TELLEZ,ALBERTO,
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: TELLEZ,ALBERTO
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 01/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 20:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:10

SUBJECT'S LAST NAME: DUARTE

DATE/TIME OF TEST: 01/25/2016 21:44:07

Diagnostics Check	OK	21:42	
Air Blank	0.000	21:43	
Control Test	0.074	21:43	Control Outside Tolerance
Air Blank	0.000	21:44	

Data Download Date/Time: 02/04/2016 11:24

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VARGAS,D,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: VARGAS,D
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 01/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 23:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:58

SUBJECT'S LAST NAME: GONZALEZ PADRON

DATE/TIME OF TEST: 01/28/2016 00:25:31

Diagnostics Check		OK	00:18
Air Blank		0.000	00:19
Control Test		0.078	00:19
Air Blank		0.000	00:20
Subject Sample	# 1	0.176	00:20
Breath Volume		2.687	00:20
Air Blank		0.000	00:21
Air Blank		0.000	00:23
Subject Sample	# 2	0.171	00:23
Breath Volume		2.312	00:23
Air Blank		0.000	00:24
Control Test		0.077	00:24
Air Blank		0.000	00:25
Diagnostics Check		OK	00:25

Data Download Date/Time: 02/04/2016 11:24

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 01/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 02:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:22

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 01/28/2016 03:49:37

Diagnostics Check		OK	03:42
Air Blank		0.000	03:43
Control Test		0.079	03:43
Air Blank		0.000	03:43
Subject Sample	# 1	0.164	03:44
Breath Volume		1.972	03:44
Air Blank		0.000	03:45
Air Blank		0.000	03:47
Subject Sample	# 2	0.177	03:47
Breath Volume		1.609	03:47
Air Blank		0.000	03:48
Control Test		0.078	03:49
Air Blank		0.000	03:49
Diagnostics Check		OK	03:49

Data Download Date/Time: 02/04/2016 11:24

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARDELAS,MAYLEE,
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: BARDELAS,MAYLEE
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 01/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 17:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:45

SUBJECT'S LAST NAME: BLANCO-MARTINEZ

DATE/TIME OF TEST: 01/28/2016 18:08:44

Diagnostics Check		OK	18:05	
Air Blank		0.000	18:06	
Control Test		0.078	18:06	
Air Blank		0.000	18:07	
Subject Sample	# 1		18:07	Subject Test Refused
Breath Volume		0.000	18:07	
Air Blank		0.000	18:07	
Control Test		0.077	18:08	
Air Blank		0.000	18:08	
Diagnostics Check		OK	18:08	

Data Download Date/Time: 02/04/2016 11:24

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 01/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 02:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: MEDINA

DATE/TIME OF TEST: 01/29/2016 03:34:34

Diagnostics Check		OK	03:26
Air Blank		0.000	03:26
Control Test		0.078	03:27
Air Blank		0.000	03:27
Subject Sample	# 1	0.180	03:29
Breath Volume		1.976	03:29
Air Blank		0.000	03:30
Air Blank		0.000	03:32
Subject Sample	# 2	0.174	03:32
Breath Volume		1.117	03:32
Air Blank		0.000	03:33
Control Test		0.077	03:33
Air Blank		0.000	03:34
Diagnostics Check		OK	03:34

Data Download Date/Time: 02/04/2016 11:24

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JIMENEZ,OSIEL,
BREATH TEST OPERATOR AGENCY: Hialeah Gardens PD
ARRESTING OFFICER: JIMENEZ,OSIEL
ARRESTING OFFICER AGENCY: Hialeah Gardens PD
LAST AGENCY INSPECTION: 01/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 21:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:38

SUBJECT'S LAST NAME: LOPEZTOPPE

DATE/TIME OF TEST: 01/29/2016 22:54:52

Diagnostics Check		OK	22:42	
Air Blank		0.000	22:42	
Control Test		0.079	22:42	
Air Blank		0.000	22:43	
Subject Sample	# 1	0.240	22:44	No .020 agreement
Breath Volume		2.332	22:44	
Air Blank		0.000	22:45	
Air Blank		0.000	22:46	
Subject Sample	# 2	0.263	22:47	No .020 agreement
Breath Volume		3.265	22:47	
Air Blank		0.000	22:48	
Air Blank		0.000	22:50	
Subject Sample	# 3		22:53	No Sample Provided
Breath Volume		0.000	22:53	
Air Blank		0.000	22:53	
Control Test		0.076	22:54	
Air Blank		0.000	22:54	
Diagnostics Check		OK	22:54	

Data Download Date/Time: 02/04/2016 11:24

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TELLEZ,ALBERTO,
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: TELLEZ,ALBERTO
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 01/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 22:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 01/29/2016 23:09:58

Diagnostics Check		OK	23:07	
Air Blank		0.000	23:07	
Control Test		0.078	23:07	
Air Blank		0.000	23:08	
Subject Sample	# 1		23:08	Subject Test Refused
Breath Volume		0.000	23:08	
Air Blank		0.000	23:09	
Control Test		0.078	23:09	
Air Blank		0.000	23:09	
Diagnostics Check		OK	23:09	

Data Download Date/Time: 02/04/2016 11:24

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,DAVID,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: CELLA,DOMENICO
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 01/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 00:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:38

SUBJECT'S LAST NAME: ESCOBAR

DATE/TIME OF TEST: 01/30/2016 01:08:08

Diagnostics Check		OK	01:01
Air Blank		0.000	01:02
Control Test		0.079	01:02
Air Blank		0.000	01:02
Subject Sample	# 1	0.163	01:03
Breath Volume		1.914	01:03
Air Blank		0.000	01:04
Air Blank		0.000	01:06
Subject Sample	# 2	0.161	01:06
Breath Volume		2.093	01:06
Air Blank		0.000	01:07
Control Test		0.076	01:07
Air Blank		0.000	01:08
Diagnostics Check		OK	01:08

Data Download Date/Time: 02/04/2016 11:24

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 01/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 04:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:02

SUBJECT'S LAST NAME: QUINTANILLA

DATE/TIME OF TEST: 01/30/2016 05:52:57

Diagnostics Check		OK	05:45
Air Blank		0.000	05:45
Control Test		0.079	05:46
Air Blank		0.000	05:46
Subject Sample	# 1	0.141	05:47
Breath Volume		1.531	05:47
Air Blank		0.000	05:48
Air Blank		0.000	05:50
Subject Sample	# 2	0.140	05:51
Breath Volume		1.769	05:51
Air Blank		0.000	05:52
Control Test		0.078	05:52
Air Blank		0.000	05:52
Diagnostics Check		OK	05:52

Data Download Date/Time: 02/04/2016 11:24

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CELLA,DOMENICO,G
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: TORRES,DAVID
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 01/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 11:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:03

SUBJECT'S LAST NAME: TORRES

DATE/TIME OF TEST: 01/31/2016 01:33:22

Diagnostics Check		OK	01:26
Air Blank		0.000	01:26
Control Test		0.078	01:27
Air Blank		0.000	01:27
Subject Sample	# 1	0.227	01:28
Breath Volume		1.507	01:28
Air Blank		0.000	01:29
Air Blank		0.000	01:30
Subject Sample	# 2	0.225	01:31
Breath Volume		1.890	01:31
Air Blank		0.000	01:32
Control Test		0.076	01:32
Air Blank		0.000	01:33
Diagnostics Check		OK	01:33

Data Download Date/Time: 02/04/2016 11:24

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: VARGAS,DAEL
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 01/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 01:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 01/31/2016 03:09:03

Diagnostics Check		OK	03:06	
Air Blank		0.000	03:06	
Control Test		0.079	03:06	
Air Blank		0.000	03:07	
Subject Sample	# 1		03:07	Subject Test Refused
Breath Volume		0.000	03:07	
Air Blank		0.000	03:08	
Control Test		0.078	03:08	
Air Blank		0.000	03:08	
Diagnostics Check		OK	03:09	

Data Download Date/Time: 02/04/2016 11:24

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 01/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:02

SUBJECT'S LAST NAME: FELDER

DATE/TIME OF TEST: 01/31/2016 03:39:46

Diagnostics Check		OK	03:32
Air Blank		0.000	03:33
Control Test		0.078	03:33
Air Blank		0.000	03:34
Subject Sample	# 1	0.133	03:34
Breath Volume		2.171	03:34
Air Blank		0.000	03:35
Air Blank		0.000	03:37
Subject Sample	# 2	0.137	03:38
Breath Volume		1.804	03:38
Air Blank		0.000	03:38
Control Test		0.079	03:39
Air Blank		0.000	03:39
Diagnostics Check		OK	03:39

Data Download Date/Time: 02/04/2016 11:24

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIANCHI,MARIO,D
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: BIANCHI,MARIO
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 01/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 15:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:51

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 01/05/2016 16:21:54

Diagnostics Check		OK	16:12
Air Blank		0.000	16:12
Control Test		0.081	16:12
Air Blank		0.000	16:13
Subject Sample	# 1	0.227	16:16
Breath Volume		2.382	16:16
Air Blank		0.000	16:17
Air Blank		0.000	16:18
Subject Sample	# 2	0.230	16:20
Breath Volume		1.898	16:20
Air Blank		0.000	16:20
Control Test		0.080	16:21
Air Blank		0.000	16:21
Diagnostics Check		OK	16:21

Data Download Date/Time: 02/04/2016 11:25

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAYARD,CLIFFORD,P
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: CAYARD,CLIFFORD
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 01/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 02:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: RODRIGUEZ TEJON

DATE/TIME OF TEST: 01/09/2016 03:10:34

Diagnostics Check		OK	03:03
Air Blank		0.000	03:03
Control Test		0.082	03:04
Air Blank		0.000	03:04
Subject Sample	# 1	0.249	03:05
Breath Volume		1.339	03:05
Air Blank		0.000	03:06
Air Blank		0.000	03:07
Subject Sample	# 2	0.246	03:08
Breath Volume		1.644	03:08
Air Blank		0.000	03:09
Control Test		0.083	03:09
Air Blank		0.000	03:10
Diagnostics Check		OK	03:10

Data Download Date/Time: 02/04/2016 11:25

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIANCHI,MARIO,D
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: BIANCHI,MARIO
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 01/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 20:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:51

SUBJECT'S LAST NAME: GOMEZOXLAJ

DATE/TIME OF TEST: 01/10/2016 21:28:17

Diagnostics Check		OK	21:21
Air Blank		0.000	21:21
Control Test		0.081	21:21
Air Blank		0.000	21:22
Subject Sample	# 1	0.268	21:23
Breath Volume		2.949	21:23
Air Blank		0.000	21:24
Air Blank		0.000	21:25
Subject Sample	# 2	0.269	21:26
Breath Volume		2.785	21:26
Air Blank		0.000	21:27
Control Test		0.081	21:27
Air Blank		0.000	21:28
Diagnostics Check		OK	21:28

Data Download Date/Time: 02/04/2016 11:25

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRANCO,GREGORY,R
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: FRANCO,GREGORY
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 01/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:04

SUBJECT'S LAST NAME: FRANCOIS

DATE/TIME OF TEST: 01/16/2016 03:34:48

Diagnostics Check		OK	03:27
Air Blank		0.000	03:27
Control Test		0.083	03:28
Air Blank		0.000	03:28
Subject Sample	# 1	0.117	03:29
Breath Volume		1.730	03:29
Air Blank		0.000	03:30
Air Blank		0.000	03:32
Subject Sample	# 2	0.117	03:33
Breath Volume		1.750	03:33
Air Blank		0.000	03:33
Control Test		0.083	03:34
Air Blank		0.000	03:34
Diagnostics Check		OK	03:34

Data Download Date/Time: 02/04/2016 11:25

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAYARD,CLIFFORD,P
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: CAYARD,CLIFFORD
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 01/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 03:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:04

SUBJECT'S LAST NAME: CITRON

DATE/TIME OF TEST: 01/17/2016 03:41:21

Diagnostics Check		OK	03:34	
Air Blank		0.000	03:34	
Control Test		0.082	03:35	
Air Blank		0.000	03:35	
Subject Sample	# 1	0.176	03:36	
Breath Volume		2.964	03:36	
Air Blank		0.000	03:37	
Air Blank		0.000	03:39	
Subject Sample	# 2	0.176	03:39	
Breath Volume		3.292	03:39	
Air Blank		0.000	03:40	
Control Test		0.080	03:40	
Air Blank			03:40	RFI Detect
Air Blank		0.000	03:41	

Data Download Date/Time: 02/04/2016 11:25

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAYARD,CLIFFORD,P
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: CAYARD,CLIFFORD
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 01/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 00:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 01/20/2016 00:54:22

Diagnostics Check		OK	00:47
Air Blank		0.000	00:48
Control Test		0.082	00:48
Air Blank		0.000	00:48
Subject Sample	# 1	0.220	00:49
Breath Volume		2.562	00:49
Air Blank		0.000	00:50
Air Blank		0.000	00:52
Subject Sample	# 2	0.219	00:52
Breath Volume		2.062	00:52
Air Blank		0.000	00:53
Control Test		0.082	00:53
Air Blank		0.000	00:54
Diagnostics Check		OK	00:54

Data Download Date/Time: 02/04/2016 11:25

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOPEZ,ERICK,A
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: MANTILLA,FRANCISCO
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 01/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 11:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 11:10

SUBJECT'S LAST NAME: OSCEOLA

DATE/TIME OF TEST: 01/21/2016 12:14:03

Diagnostics Check		OK	12:07
Air Blank		0.000	12:08
Control Test		0.082	12:08
Air Blank		0.000	12:08
Subject Sample	# 1	0.000	12:09
Breath Volume		1.371	12:09
Air Blank		0.000	12:10
Air Blank		0.000	12:12
Subject Sample	# 2	0.000	12:12
Breath Volume		1.449	12:12
Air Blank		0.000	12:13
Control Test		0.081	12:13
Air Blank		0.000	12:13
Diagnostics Check		OK	12:14

Data Download Date/Time: 02/04/2016 11:25

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIANCHI,MARIO,D
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: BIANCHI,MARIO
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 01/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 20:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:14

SUBJECT'S LAST NAME: SANCHEZ JIMENEZ

DATE/TIME OF TEST: 01/23/2016 21:38:50

Diagnostics Check	OK	21:37	
Air Blank	0.000	21:37	
Control Test	0.081	21:38	
Air Blank		21:38	RFI Detect
Air Blank	0.000	21:38	

Data Download Date/Time: 02/04/2016 11:25

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIANCHI,MARIO,D
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: BIANCHI,MARIO
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 01/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 20:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:14

SUBJECT'S LAST NAME: SANCHEZ JIMENEZ

DATE/TIME OF TEST: 01/23/2016 21:52:08

Diagnostics Check		OK	21:43
Air Blank		0.000	21:43
Control Test		0.082	21:43
Air Blank		0.000	21:44
Subject Sample	# 1	0.202	21:45
Breath Volume		1.640	21:45
Air Blank		0.000	21:45
Air Blank		0.000	21:47
Subject Sample	# 2	0.200	21:50
Breath Volume		2.226	21:50
Air Blank		0.000	21:51
Control Test		0.083	21:51
Air Blank		0.000	21:52
Diagnostics Check		OK	21:52

Data Download Date/Time: 02/04/2016 11:25

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAYARD,CLIFFORD,P
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: CAYARD,CLIFFORD
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 01/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 22:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: BYNUM III

DATE/TIME OF TEST: 01/24/2016 23:02:12

Diagnostics Check	OK	23:01	
Air Blank		23:01	RFI Detect
Air Blank	0.000	23:02	

Data Download Date/Time: 02/04/2016 11:25

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAYARD,CLIFFORD,P
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: CAYARD,CLIFFORD
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 01/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 22:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: BYNUM III

DATE/TIME OF TEST: 01/24/2016 23:11:41

Diagnostics Check		OK	23:03
Air Blank		0.000	23:04
Control Test		0.083	23:04
Air Blank		0.000	23:05
Subject Sample	# 1	0.284	23:06
Breath Volume		1.714	23:06
Air Blank		0.000	23:06
Air Blank		0.000	23:08
Subject Sample	# 2	0.277	23:09
Breath Volume		2.691	23:09
Air Blank		0.000	23:10
Control Test		0.082	23:11
Air Blank		0.000	23:11
Diagnostics Check		OK	23:11

Data Download Date/Time: 02/04/2016 11:25

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELLON,RICHARD,G
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: BELLON,RICHARD
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 01/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 04:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:28

SUBJECT'S LAST NAME: OLMOS

DATE/TIME OF TEST: 01/27/2016 05:55:40

Diagnostics Check		OK	05:49	
Air Blank		0.000	05:49	
Control Test		0.081	05:49	
Air Blank		0.000	05:50	
Subject Sample	# 1	0.245	05:51	
Breath Volume		2.773	05:51	
Air Blank		0.000	05:51	
Air Blank		0.000	05:53	
Subject Sample	# 2	0.246	05:53	
Breath Volume		3.710	05:53	
Air Blank		0.000	05:54	
Control Test		0.088	05:55	Control Outside Tolerance
Air Blank		0.000	05:55	

Data Download Date/Time: 02/04/2016 11:25

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SLIMAK,M,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: SLIMAK,M
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 01/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 22:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:55

SUBJECT'S LAST NAME: MELGAR-OROZCO

DATE/TIME OF TEST: 01/27/2016 23:20:07

Diagnostics Check		OK	23:17	
Air Blank		0.000	23:17	
Control Test		0.081	23:17	
Air Blank		0.000	23:18	
Subject Sample	# 1		23:18	RFI Detect
Breath Volume		1.187	23:18	
Air Blank		0.000	23:19	
Control Test		0.074	23:19	Control Outside Tolerance
Air Blank		0.000	23:20	

Data Download Date/Time: 02/04/2016 11:26

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SLIMAK,M,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: SLIMAK,M
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 01/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 22:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:55

SUBJECT'S LAST NAME: MELGAR-OROZCO

DATE/TIME OF TEST: 01/27/2016 23:29:17

Diagnostics Check		OK	23:22
Air Blank		0.000	23:23
Control Test		0.082	23:23
Air Blank		0.000	23:23
Subject Sample	# 1	0.197	23:24
Breath Volume		2.480	23:24
Air Blank		0.000	23:25
Air Blank		0.000	23:27
Subject Sample	# 2	0.194	23:27
Breath Volume		2.003	23:27
Air Blank		0.000	23:28
Control Test		0.081	23:28
Air Blank		0.000	23:29
Diagnostics Check		OK	23:29

Data Download Date/Time: 02/04/2016 11:26

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REEVE,YAARA,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: REEVE,YAARA
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 01/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 22:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:58

SUBJECT'S LAST NAME: RODRIGUE

DATE/TIME OF TEST: 01/09/2016 23:34:40

Diagnostics Check		OK	23:31	
Air Blank		0.000	23:32	
Control Test		0.080	23:32	
Air Blank		0.000	23:33	
Subject Sample	# 1		23:33	Subject Test Refused
Breath Volume		0.000	23:33	
Air Blank		0.000	23:33	
Control Test		0.080	23:34	
Air Blank		0.000	23:34	
Diagnostics Check		OK	23:34	

Data Download Date/Time: 01/15/2016 11:35

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLORENIO,RAFAEL,E
BREATH TEST OPERATOR AGENCY: North Miami Beach PD
ARRESTING OFFICER: FLORENCIO,RAFAEL
ARRESTING OFFICER AGENCY: North Miami Beach PD
LAST AGENCY INSPECTION: 12/28/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1878845
DRY GAS EXPIRY DATE: 06/22/2018
VIOLATION CODE: DUI
ARREST TIME: 03:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:22

SUBJECT'S LAST NAME: CLIFFORD

DATE/TIME OF TEST: 01/01/2016 03:53:50

Diagnostics Check		OK	03:47
Air Blank		0.000	03:47
Control Test		0.079	03:48
Air Blank		0.000	03:48
Subject Sample	# 1	0.177	03:49
Breath Volume		1.347	03:49
Air Blank		0.000	03:50
Air Blank		0.000	03:51
Subject Sample	# 2	0.182	03:52
Breath Volume		1.402	03:52
Air Blank		0.000	03:52
Control Test		0.078	03:53
Air Blank		0.000	03:53
Diagnostics Check		OK	03:53

Data Download Date/Time: 01/28/2016 22:44

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAYA,JOSE,
BREATH TEST OPERATOR AGENCY: North Miami Beach PD
ARRESTING OFFICER: ACEVEDO,MICHAEL
ARRESTING OFFICER AGENCY: North Miami Beach PD
LAST AGENCY INSPECTION: 12/28/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1878845
DRY GAS EXPIRY DATE: 06/22/2018
VIOLATION CODE: DUI
ARREST TIME: 01:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:44

SUBJECT'S LAST NAME: LOUIS

DATE/TIME OF TEST: 01/03/2016 02:11:49

Diagnostics Check		OK	02:04
Air Blank		0.000	02:05
Control Test		0.079	02:05
Air Blank		0.000	02:05
Subject Sample	# 1	0.118	02:06
Breath Volume		1.156	02:06
Air Blank		0.000	02:07
Air Blank		0.000	02:08
Subject Sample	# 2	0.122	02:10
Breath Volume		1.234	02:10
Air Blank		0.000	02:10
Control Test		0.077	02:11
Air Blank		0.000	02:11
Diagnostics Check		OK	02:11

Data Download Date/Time: 01/28/2016 22:44

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLORENCIO,RAFAEL,E
BREATH TEST OPERATOR AGENCY: North Miami Beach PD
ARRESTING OFFICER: FLORENCIO,RAFAEL
ARRESTING OFFICER AGENCY: North Miami Beach PD
LAST AGENCY INSPECTION: 12/28/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1878845
DRY GAS EXPIRY DATE: 06/22/2018
VIOLATION CODE: DUI
ARREST TIME: 04:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:29

SUBJECT'S LAST NAME: VARGAS

DATE/TIME OF TEST: 01/07/2016 04:56:46

Diagnostics Check		OK	04:49
Air Blank		0.000	04:49
Control Test		0.080	04:50
Air Blank		0.000	04:50
Subject Sample	# 1	0.117	04:51
Breath Volume		2.128	04:51
Air Blank		0.000	04:52
Air Blank		0.000	04:54
Subject Sample	# 2	0.115	04:55
Breath Volume		2.117	04:55
Air Blank		0.000	04:55
Control Test		0.078	04:56
Air Blank		0.000	04:56
Diagnostics Check		OK	04:56

Data Download Date/Time: 01/28/2016 22:44

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAKER,KENNETH,
BREATH TEST OPERATOR AGENCY: North Miami Beach PD
ARRESTING OFFICER: SIMON,SAM
ARRESTING OFFICER AGENCY: North Miami Beach PD
LAST AGENCY INSPECTION: 12/28/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1878845
DRY GAS EXPIRY DATE: 06/22/2018
VIOLATION CODE: DUI
ARREST TIME: 01:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: MEJIA

DATE/TIME OF TEST: 01/13/2016 02:35:29

Diagnostics Check		OK	02:27
Air Blank		0.000	02:27
Control Test		0.079	02:28
Air Blank		0.000	02:28
Subject Sample	# 1	0.196	02:29
Breath Volume		1.250	02:29
Air Blank		0.000	02:30
Air Blank		0.000	02:32
Subject Sample	# 2	0.199	02:33
Breath Volume		1.312	02:33
Air Blank		0.000	02:34
Control Test		0.078	02:34
Air Blank		0.000	02:35
Diagnostics Check		OK	02:35

Data Download Date/Time: 01/28/2016 22:44

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAYA,JOSE,
BREATH TEST OPERATOR AGENCY: North Miami Beach PD
ARRESTING OFFICER: MAYA,JOSE
ARRESTING OFFICER AGENCY: North Miami Beach PD
LAST AGENCY INSPECTION: 12/28/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1878845
DRY GAS EXPIRY DATE: 06/22/2018
VIOLATION CODE: DUI
ARREST TIME: 02:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:58

SUBJECT'S LAST NAME: MENDEZ

DATE/TIME OF TEST: 01/14/2016 03:24:58

Diagnostics Check		OK	03:18
Air Blank		0.000	03:19
Control Test		0.080	03:19
Air Blank		0.000	03:19
Subject Sample	# 1	0.095	03:20
Breath Volume		1.570	03:20
Air Blank		0.000	03:21
Air Blank		0.000	03:23
Subject Sample	# 2	0.090	03:23
Breath Volume		1.320	03:23
Air Blank		0.000	03:24
Control Test		0.078	03:24
Air Blank		0.000	03:24
Diagnostics Check		OK	03:24

Data Download Date/Time: 01/28/2016 22:44

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAKER,KENNETH,
BREATH TEST OPERATOR AGENCY: North Miami Beach PD
ARRESTING OFFICER: FARINI,K
ARRESTING OFFICER AGENCY: North Miami Beach PD
LAST AGENCY INSPECTION: 12/28/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1878845
DRY GAS EXPIRY DATE: 06/22/2018
VIOLATION CODE: DUI
ARREST TIME: 03:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: LATTIMORE

DATE/TIME OF TEST: 01/14/2016 03:57:01

Diagnostics Check		OK	03:49
Air Blank		0.000	03:50
Control Test		0.079	03:50
Air Blank		0.000	03:50
Subject Sample	# 1	0.068	03:52
Breath Volume		1.562	03:52
Air Blank		0.000	03:52
Air Blank		0.000	03:54
Subject Sample	# 2	0.065	03:55
Breath Volume		1.527	03:55
Air Blank		0.000	03:56
Control Test		0.079	03:56
Air Blank		0.000	03:56
Diagnostics Check		OK	03:57

Data Download Date/Time: 01/28/2016 22:44

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAYA,JOSE,
BREATH TEST OPERATOR AGENCY: North Miami Beach PD
ARRESTING OFFICER: KELNER,D
ARRESTING OFFICER AGENCY: North Miami Beach PD
LAST AGENCY INSPECTION: 12/28/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1878845
DRY GAS EXPIRY DATE: 06/22/2018
VIOLATION CODE: DUI
ARREST TIME: 23:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:33

SUBJECT'S LAST NAME: CRUZ

DATE/TIME OF TEST: 01/15/2016 00:01:25

Diagnostics Check		OK	23:53
Air Blank		0.000	23:53
Control Test		0.079	23:54
Air Blank		0.000	23:54
Subject Sample	# 1	0.200	23:55
Breath Volume		1.234	23:55
Air Blank		0.000	23:56
Air Blank		0.000	23:58
Subject Sample	# 2	0.187	23:59
Breath Volume		1.503	23:59
Air Blank		0.000	00:00
Control Test		0.076	00:00
Air Blank		0.000	00:01
Diagnostics Check		OK	00:01

Data Download Date/Time: 01/28/2016 22:44

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAYA,JOSE,
BREATH TEST OPERATOR AGENCY: North Miami Beach PD
ARRESTING OFFICER: BONY,M
ARRESTING OFFICER AGENCY: North Miami PD
LAST AGENCY INSPECTION: 12/28/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1878845
DRY GAS EXPIRY DATE: 06/22/2018
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: GARCIA DE LA NOCEDA

DATE/TIME OF TEST: 01/16/2016 03:23:37

Diagnostics Check		OK	03:15
Air Blank		0.000	03:15
Control Test		0.080	03:16
Air Blank		0.000	03:16
Subject Sample	# 1	0.172	03:18
Breath Volume		1.398	03:18
Air Blank		0.000	03:19
Air Blank		0.000	03:20
Subject Sample	# 2	0.189	03:21
Breath Volume		1.230	03:21
Air Blank		0.000	03:22
Control Test		0.077	03:23
Air Blank		0.000	03:23
Diagnostics Check		OK	03:23

Data Download Date/Time: 01/28/2016 22:44

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAYA,JOSE,
BREATH TEST OPERATOR AGENCY: North Miami Beach PD
ARRESTING OFFICER: KELLNER,D
ARRESTING OFFICER AGENCY: North Miami Beach PD
LAST AGENCY INSPECTION: 12/28/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1878845
DRY GAS EXPIRY DATE: 06/22/2018
VIOLATION CODE: DUI
ARREST TIME: 03:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:14

SUBJECT'S LAST NAME: MEISTER

DATE/TIME OF TEST: 01/16/2016 03:44:15

Diagnostics Check		OK	03:36
Air Blank		0.000	03:36
Control Test		0.079	03:37
Air Blank		0.000	03:37
Subject Sample	# 1	0.137	03:39
Breath Volume		1.921	03:39
Air Blank		0.000	03:39
Air Blank		0.000	03:41
Subject Sample	# 2	0.142	03:42
Breath Volume		1.343	03:42
Air Blank		0.000	03:43
Control Test		0.077	03:43
Air Blank		0.000	03:44
Diagnostics Check		OK	03:44

Data Download Date/Time: 01/28/2016 22:44

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAKER,KENNETH,
BREATH TEST OPERATOR AGENCY: North Miami Beach PD
ARRESTING OFFICER: AZEVEDO,M
ARRESTING OFFICER AGENCY: North Miami Beach PD
LAST AGENCY INSPECTION: 12/28/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1878845
DRY GAS EXPIRY DATE: 06/22/2018
VIOLATION CODE: DUI
ARREST TIME: 02:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:54

SUBJECT'S LAST NAME: SALMERON ULLOA

DATE/TIME OF TEST: 01/21/2016 03:23:59

Diagnostics Check		OK	03:15
Air Blank		0.000	03:15
Control Test		0.080	03:15
Air Blank		0.000	03:16
Subject Sample	# 1	0.192	03:18
Breath Volume		1.707	03:18
Air Blank		0.000	03:19
Air Blank		0.000	03:21
Subject Sample	# 2	0.204	03:22
Breath Volume		1.460	03:22
Air Blank		0.000	03:23
Control Test		0.078	03:23
Air Blank		0.000	03:23
Diagnostics Check		OK	03:23

Data Download Date/Time: 01/28/2016 22:44

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAYA,JOSE,
BREATH TEST OPERATOR AGENCY: North Miami Beach PD
ARRESTING OFFICER: MAYA,JOSE
ARRESTING OFFICER AGENCY: North Miami Beach PD
LAST AGENCY INSPECTION: 12/28/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1878845
DRY GAS EXPIRY DATE: 06/22/2018
VIOLATION CODE: DUI
ARREST TIME: 03:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:39

SUBJECT'S LAST NAME: HANLEY

DATE/TIME OF TEST: 01/24/2016 04:02:19

Diagnostics Check		OK	03:59	
Air Blank		0.000	03:59	
Control Test		0.080	04:00	
Air Blank		0.000	04:00	
Subject Sample	# 1		04:00	Subject Test Refused
Breath Volume		0.000	04:00	
Air Blank		0.000	04:01	
Control Test		0.081	04:01	
Air Blank		0.000	04:02	
Diagnostics Check		OK	04:02	

Data Download Date/Time: 01/28/2016 22:44

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAKER,KENNETH,
BREATH TEST OPERATOR AGENCY: North Miami Beach PD
ARRESTING OFFICER: AZEVEDO,M
ARRESTING OFFICER AGENCY: North Miami Beach PD
LAST AGENCY INSPECTION: 12/28/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1878845
DRY GAS EXPIRY DATE: 06/22/2018
VIOLATION CODE: DUI
ARREST TIME: 04:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:27

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 01/26/2016 04:52:42

Diagnostics Check		OK	04:49	
Air Blank		0.000	04:49	
Control Test		0.080	04:49	
Air Blank		0.000	04:50	
Subject Sample	# 1		04:51	Subject Test Refused
Breath Volume		0.000	04:51	
Air Blank		0.000	04:51	
Control Test		0.079	04:52	
Air Blank		0.000	04:52	
Diagnostics Check		OK	04:52	

Data Download Date/Time: 01/28/2016 22:44

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAKER,KENNETH,
BREATH TEST OPERATOR AGENCY: North Miami Beach PD
ARRESTING OFFICER: AZEVEDO,M
ARRESTING OFFICER AGENCY: North Miami Beach PD
LAST AGENCY INSPECTION: 12/28/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1878845
DRY GAS EXPIRY DATE: 06/22/2018
VIOLATION CODE: DUI
ARREST TIME: 03:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: BRYAN

DATE/TIME OF TEST: 01/27/2016 03:59:33

Diagnostics Check		OK	03:56	
Air Blank		0.000	03:56	
Control Test		0.079	03:57	
Air Blank		0.000	03:57	
Subject Sample	# 1		03:58	Subject Test Refused
Breath Volume		0.000	03:58	
Air Blank		0.000	03:58	
Control Test		0.079	03:58	
Air Blank		0.000	03:59	
Diagnostics Check		OK	03:59	

Data Download Date/Time: 01/28/2016 22:44

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLORENCIO,RAFAEL,E
BREATH TEST OPERATOR AGENCY: North Miami Beach PD
ARRESTING OFFICER: MAYA,JOSE
ARRESTING OFFICER AGENCY: North Miami Beach PD
LAST AGENCY INSPECTION: 01/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1878845
DRY GAS EXPIRY DATE: 06/22/2018
VIOLATION CODE: DUI
ARREST TIME: 02:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:06

SUBJECT'S LAST NAME: CAMACHO

DATE/TIME OF TEST: 01/29/2016 03:29:39

Diagnostics Check		OK	03:26	
Air Blank		0.000	03:27	
Control Test		0.080	03:27	
Air Blank		0.000	03:28	
Subject Sample	# 1		03:28	Subject Test Refused
Breath Volume		0.000	03:28	
Air Blank		0.000	03:28	
Control Test		0.080	03:29	
Air Blank		0.000	03:29	
Diagnostics Check		OK	03:29	

Data Download Date/Time: 02/28/2016 08:51

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLORENCIO,RAFAEL,E
BREATH TEST OPERATOR AGENCY: North Miami Beach PD
ARRESTING OFFICER: FLORENCIO,RAFAEL
ARRESTING OFFICER AGENCY: North Miami Beach PD
LAST AGENCY INSPECTION: 01/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1878845
DRY GAS EXPIRY DATE: 06/22/2018
VIOLATION CODE: DUI
ARREST TIME: 04:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:15

SUBJECT'S LAST NAME: DEL POZO

DATE/TIME OF TEST: 01/29/2016 04:43:10

Diagnostics Check		OK	04:40	
Air Blank		0.000	04:40	
Control Test		0.080	04:41	
Air Blank		0.000	04:41	
Subject Sample	# 1		04:41	Subject Test Refused
Breath Volume		0.000	04:41	
Air Blank		0.000	04:42	
Control Test		0.079	04:42	
Air Blank		0.000	04:43	
Diagnostics Check		OK	04:43	

Data Download Date/Time: 02/28/2016 08:51

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAYA,JOSE,
BREATH TEST OPERATOR AGENCY: North Miami Beach PD
ARRESTING OFFICER: MAYA,JOSE
ARRESTING OFFICER AGENCY: North Miami Beach PD
LAST AGENCY INSPECTION: 01/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1878845
DRY GAS EXPIRY DATE: 06/22/2018
VIOLATION CODE: DUI
ARREST TIME: 01:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:19

SUBJECT'S LAST NAME: MENDOZA UGARTE

DATE/TIME OF TEST: 01/30/2016 01:47:36

Diagnostics Check		OK	01:39
Air Blank		0.000	01:40
Control Test		0.080	01:40
Air Blank		0.000	01:40
Subject Sample	# 1	0.112	01:43
Breath Volume		1.308	01:43
Air Blank		0.000	01:43
Air Blank		0.000	01:45
Subject Sample	# 2	0.124	01:46
Breath Volume		1.261	01:46
Air Blank		0.000	01:46
Control Test		0.077	01:47
Air Blank		0.000	01:47
Diagnostics Check		OK	01:47

Data Download Date/Time: 02/28/2016 08:51

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCKENZIE,DEPUTY,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MCKENZIE,DEPUTY
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/31/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 13:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:46

SUBJECT'S LAST NAME: DANCU

DATE/TIME OF TEST: 01/12/2016 14:13:55

Diagnostics Check		OK	14:06
Air Blank		0.000	14:07
Control Test		0.081	14:07
Air Blank		0.000	14:08
Subject Sample	# 1	0.000	14:09
Breath Volume		1.226	14:09
Air Blank		0.000	14:10
Air Blank		0.000	14:12
Subject Sample	# 2	0.000	14:12
Breath Volume		1.847	14:12
Air Blank		0.000	14:13
Control Test		0.081	14:13
Air Blank		0.000	14:13
Diagnostics Check		OK	14:13

Data Download Date/Time: 01/30/2016 20:08

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000892

INSTRUMENT REGISTERED TO: FWC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RANGEL,STEVEN,A
BREATH TEST OPERATOR AGENCY: Naval Lab Pana City
ARRESTING OFFICER: JACKSON,KARL
ARRESTING OFFICER AGENCY: Naval Lab Pana City
LAST AGENCY INSPECTION: 12/16/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28214080A1
DRY GAS EXPIRY DATE: 11/05/2016
VIOLATION CODE: DUI
ARREST TIME: 09:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:15

SUBJECT'S LAST NAME: PRITCHARD

DATE/TIME OF TEST: 01/18/2016 09:53:26

Diagnostics Check		OK	09:47
Air Blank		0.000	09:47
Control Test		0.079	09:47
Air Blank		0.000	09:48
Subject Sample	# 1	0.036	09:48
Breath Volume		3.757	09:48
Air Blank		0.000	09:49
Air Blank		0.000	09:51
Subject Sample	# 2	0.032	09:51
Breath Volume		3.714	09:51
Air Blank		0.000	09:52
Control Test		0.077	09:52
Air Blank		0.000	09:53
Diagnostics Check		OK	09:53

Data Download Date/Time: 01/21/2016 13:21

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000901

INSTRUMENT REGISTERED TO: FFWCC TAMPA
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHEFANO,RICHARD,A
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)
ARRESTING OFFICER: MARTINEZ,ALEX
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 01/18/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03414080A3
DRY GAS EXPIRY DATE: 03/01/2016
VIOLATION CODE: BUI
ARREST TIME: 15:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:16

SUBJECT'S LAST NAME: FERNANDEZ

DATE/TIME OF TEST: 01/30/2016 16:01:12

Diagnostics Check		OK	15:54
Air Blank		0.000	15:55
Control Test		0.080	15:55
Air Blank		0.000	15:56
Subject Sample	# 1	0.051	15:56
Breath Volume		1.972	15:56
Air Blank		0.000	15:57
Air Blank		0.000	15:59
Subject Sample	# 2	0.049	15:59
Breath Volume		1.925	15:59
Air Blank		0.000	16:00
Control Test		0.080	16:00
Air Blank		0.000	16:01
Diagnostics Check		OK	16:01

Data Download Date/Time: 02/18/2016 18:36

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000901

INSTRUMENT REGISTERED TO: FFWCC TAMPA
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHEFANO,RICHARD,A
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)
ARRESTING OFFICER: SEHL,ERIN
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 01/18/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03414080A3
DRY GAS EXPIRY DATE: 03/01/2016
VIOLATION CODE: BUI
ARREST TIME: 15:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:25

SUBJECT'S LAST NAME: HUITT

DATE/TIME OF TEST: 01/30/2016 16:20:33

Diagnostics Check		OK	16:14
Air Blank		0.000	16:14
Control Test		0.079	16:15
Air Blank		0.000	16:15
Subject Sample	# 1	0.031	16:16
Breath Volume		1.972	16:16
Air Blank		0.000	16:16
Air Blank		0.000	16:18
Subject Sample	# 2	0.031	16:19
Breath Volume		3.277	16:19
Air Blank		0.000	16:19
Control Test		0.080	16:19
Air Blank		0.000	16:20
Diagnostics Check		OK	16:20

Data Download Date/Time: 02/18/2016 18:36

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000901

INSTRUMENT REGISTERED TO: FFWCC TAMPA
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROOKS,JOSEPH,P
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)
ARRESTING OFFICER: HUGHES,GEORGE
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 01/18/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03414080A3
DRY GAS EXPIRY DATE: 03/01/2016
VIOLATION CODE: BUI
ARREST TIME: 16:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:15

SUBJECT'S LAST NAME: BARR

DATE/TIME OF TEST: 01/30/2016 16:42:52

Diagnostics Check		OK	16:36
Air Blank		0.000	16:36
Control Test		0.079	16:37
Air Blank		0.000	16:37
Subject Sample	# 1	0.102	16:38
Breath Volume		3.722	16:38
Air Blank		0.000	16:38
Air Blank		0.000	16:40
Subject Sample	# 2	0.102	16:41
Breath Volume		3.480	16:41
Air Blank		0.000	16:41
Control Test		0.079	16:42
Air Blank		0.000	16:42
Diagnostics Check		OK	16:42

Data Download Date/Time: 02/18/2016 18:36

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000901

INSTRUMENT REGISTERED TO: FFWCC TAMPA
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROOKS,JOSEPH,P
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)
ARRESTING OFFICER: ALVIS,REESE
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 01/18/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03414080A3
DRY GAS EXPIRY DATE: 03/01/2016
VIOLATION CODE: BUI
ARREST TIME: 18:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:18

SUBJECT'S LAST NAME: QUEVEDO III

DATE/TIME OF TEST: 01/30/2016 18:45:24

Diagnostics Check		OK	18:39
Air Blank		0.000	18:39
Control Test		0.080	18:40
Air Blank		0.000	18:40
Subject Sample	# 1	0.103	18:41
Breath Volume		3.269	18:41
Air Blank		0.000	18:41
Air Blank		0.000	18:43
Subject Sample	# 2	0.095	18:43
Breath Volume		3.261	18:43
Air Blank		0.000	18:44
Control Test		0.080	18:44
Air Blank		0.000	18:45
Diagnostics Check		OK	18:45

Data Download Date/Time: 02/18/2016 18:36

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000901

INSTRUMENT REGISTERED TO: FFWCC TAMPA
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROOKS,JOSEPH,P
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)
ARRESTING OFFICER: BURKES,JOHN
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 01/18/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03414080A3
DRY GAS EXPIRY DATE: 03/01/2016
VIOLATION CODE: BUI
ARREST TIME: 18:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:35

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 01/30/2016 19:04:25

Diagnostics Check		OK	18:58
Air Blank		0.000	18:58
Control Test		0.080	18:58
Air Blank		0.000	18:59
Subject Sample	# 1	0.083	19:00
Breath Volume		2.285	19:00
Air Blank		0.000	19:00
Air Blank		0.000	19:02
Subject Sample	# 2	0.078	19:02
Breath Volume		2.074	19:02
Air Blank		0.000	19:03
Control Test		0.080	19:03
Air Blank		0.000	19:04
Diagnostics Check		OK	19:04

Data Download Date/Time: 02/18/2016 18:36

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000937

INSTRUMENT REGISTERED TO: SANIBEL ISLAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUIKEMA,ELIZABETH,M
BREATH TEST OPERATOR AGENCY: Sanibel PD
ARRESTING OFFICER: JAKUBOWSKI,JOHN
ARRESTING OFFICER AGENCY: Sanibel PD
LAST AGENCY INSPECTION: 12/26/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05514080A1
DRY GAS EXPIRY DATE: 04/01/2016
VIOLATION CODE: DUI
ARREST TIME: 18:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:55

SUBJECT'S LAST NAME: RODRIGUEZ GOMEZ

DATE/TIME OF TEST: 01/15/2016 19:26:30

Diagnostics Check		OK	19:21	
Air Blank		0.000	19:22	
Control Test		0.080	19:22	
Air Blank		0.000	19:23	
Subject Sample	# 1	0.169	19:25	
Breath Volume		2.550	19:25	
Air Blank			19:25	Purge Fail
Air Blank		0.000	19:26	

Data Download Date/Time: 01/22/2016 01:19

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000937

INSTRUMENT REGISTERED TO: SANIBEL ISLAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUIKEMA,ELIZABETH,M
BREATH TEST OPERATOR AGENCY: Sanibel PD
ARRESTING OFFICER: JAKUBOWSKI,JOHN
ARRESTING OFFICER AGENCY: Sanibel PD
LAST AGENCY INSPECTION: 12/26/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05514080A1
DRY GAS EXPIRY DATE: 04/01/2016
VIOLATION CODE: DUI
ARREST TIME: 18:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:55

SUBJECT'S LAST NAME: RODRIGUEZ GOMEZ

DATE/TIME OF TEST: 01/15/2016 19:39:53

Diagnostics Check		OK	19:33
Air Blank		0.000	19:33
Control Test		0.079	19:34
Air Blank		0.000	19:34
Subject Sample	# 1	0.156	19:35
Breath Volume		1.609	19:35
Air Blank		0.000	19:35
Air Blank		0.000	19:37
Subject Sample	# 2	0.153	19:38
Breath Volume		3.277	19:38
Air Blank		0.000	19:38
Control Test		0.081	19:39
Air Blank		0.000	19:39
Diagnostics Check		OK	19:39

Data Download Date/Time: 01/22/2016 01:19

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAN HELDEN,JOSEPH,G
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: RALL,THOMAS
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: RUGULO

DATE/TIME OF TEST: 01/01/2016 23:11:54

Diagnostics Check		OK	23:00
Air Blank		0.000	23:01
Control Test		0.081	23:01
Air Blank		0.000	23:01
Subject Sample	# 1	0.232	23:02
Breath Volume		1.683	23:02
Air Blank		0.000	23:03
Air Blank		0.000	23:05
Subject Sample	# 2	0.256	23:06
Breath Volume		2.757	23:06
Air Blank		0.000	23:07
Air Blank		0.000	23:09
Subject Sample	# 3	0.248	23:10
Breath Volume		2.917	23:10
Air Blank		0.000	23:10
Control Test		0.078	23:11
Air Blank		0.000	23:11
Diagnostics Check		OK	23:11

Data Download Date/Time: 01/25/2016 09:55

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAPUZZI,KRISTIN,G
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: GROSSI,BLAKE
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: BRIGGS

DATE/TIME OF TEST: 01/06/2016 03:53:08

Diagnostics Check		OK	03:46
Air Blank		0.000	03:47
Control Test		0.080	03:47
Air Blank		0.000	03:48
Subject Sample	# 1	0.095	03:48
Breath Volume		2.484	03:48
Air Blank		0.000	03:49
Air Blank		0.000	03:51
Subject Sample	# 2	0.102	03:51
Breath Volume		2.636	03:51
Air Blank		0.000	03:52
Control Test		0.079	03:52
Air Blank		0.000	03:53
Diagnostics Check		OK	03:53

Data Download Date/Time: 01/25/2016 09:55

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAPUZZI,KRISTIN,G
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: MACKERETH,TYLER
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 01/06/2016 04:04:06

Diagnostics Check		OK	03:57
Air Blank		0.000	03:58
Control Test		0.081	03:58
Air Blank		0.000	03:59
Subject Sample	# 1	0.230	03:59
Breath Volume		2.660	03:59
Air Blank		0.000	04:00
Air Blank		0.000	04:02
Subject Sample	# 2	0.240	04:02
Breath Volume		2.410	04:02
Air Blank		0.000	04:03
Control Test		0.080	04:03
Air Blank		0.000	04:04
Diagnostics Check		OK	04:04

Data Download Date/Time: 01/25/2016 09:55

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SERRATO,JOSHUA,M
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: CONNELL,NOLAN
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 11:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:45

SUBJECT'S LAST NAME: SALVADOR-JUAN

DATE/TIME OF TEST: 01/10/2016 12:17:35

Diagnostics Check		OK	12:10
Air Blank		0.000	12:11
Control Test		0.081	12:11
Air Blank		0.000	12:11
Subject Sample	# 1	0.316	12:12
Breath Volume		2.171	12:12
Air Blank		0.000	12:13
Air Blank		0.000	12:14
Subject Sample	# 2	0.322	12:15
Breath Volume		2.191	12:15
Air Blank		0.000	12:16
Control Test		0.081	12:16
Air Blank		0.000	12:17
Diagnostics Check		OK	12:17

Data Download Date/Time: 01/25/2016 09:55

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEYERS,FREDERICK,E
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: GALLAGHER,R
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 16:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:37

SUBJECT'S LAST NAME: TETREAULT

DATE/TIME OF TEST: 01/14/2016 18:22:02

Diagnostics Check		OK	18:14
Air Blank		0.000	18:15
Control Test		0.080	18:15
Air Blank		0.000	18:16
Subject Sample	# 1	0.202	18:17
Breath Volume		1.210	18:17
Air Blank		0.000	18:17
Air Blank		0.000	18:19
Subject Sample	# 2	0.206	18:20
Breath Volume		1.562	18:20
Air Blank		0.000	18:21
Control Test		0.080	18:21
Air Blank		0.000	18:21
Diagnostics Check		OK	18:22

Data Download Date/Time: 01/25/2016 09:55

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROJAS,RANDY,
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: ROJAS,RANDY
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 19:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:42

SUBJECT'S LAST NAME: THORNLEY

DATE/TIME OF TEST: 01/14/2016 21:42:18

Diagnostics Check		OK	21:35
Air Blank		0.000	21:36
Control Test		0.081	21:36
Air Blank		0.000	21:36
Subject Sample	# 1	0.111	21:37
Breath Volume		1.140	21:37
Air Blank		0.000	21:37
Air Blank		0.000	21:39
Subject Sample	# 2	0.130	21:40
Breath Volume		2.171	21:40
Air Blank		0.000	21:41
Control Test		0.082	21:41
Air Blank		0.000	21:42
Diagnostics Check		OK	21:42

Data Download Date/Time: 01/25/2016 09:55

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRASK,J,N
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: CARTER,C
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:14

SUBJECT'S LAST NAME: VAN NETTA

DATE/TIME OF TEST: 01/17/2016 03:41:00

Diagnostics Check		OK	03:34
Air Blank		0.000	03:35
Control Test		0.080	03:35
Air Blank		0.000	03:35
Subject Sample	# 1	0.182	03:36
Breath Volume		2.792	03:36
Air Blank		0.000	03:37
Air Blank		0.000	03:38
Subject Sample	# 2	0.191	03:39
Breath Volume		2.968	03:39
Air Blank		0.000	03:40
Control Test		0.080	03:40
Air Blank		0.000	03:40
Diagnostics Check		OK	03:41

Data Download Date/Time: 01/25/2016 09:55

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAN HELDEN,JOSEPH,G
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: ORO,JORGE
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: DUNCAN

DATE/TIME OF TEST: 01/18/2016 02:29:37

Diagnostics Check		OK	02:22
Air Blank		0.000	02:23
Control Test		0.080	02:23
Air Blank		0.000	02:23
Subject Sample	# 1	0.150	02:24
Breath Volume		2.472	02:24
Air Blank		0.000	02:25
Air Blank		0.000	02:27
Subject Sample	# 2	0.139	02:28
Breath Volume		2.617	02:28
Air Blank		0.000	02:28
Control Test		0.080	02:29
Air Blank		0.000	02:29
Diagnostics Check		OK	02:29

Data Download Date/Time: 01/25/2016 09:55

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STROBEL,RAYMOND,R
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: STROBEL,RAYMOND
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 08:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:05

SUBJECT'S LAST NAME: THERIAULT

DATE/TIME OF TEST: 01/22/2016 09:35:03

Diagnostics Check		OK	09:28
Air Blank		0.000	09:28
Control Test		0.080	09:29
Air Blank		0.000	09:29
Subject Sample	# 1	0.104	09:30
Breath Volume		2.214	09:30
Air Blank		0.000	09:31
Air Blank		0.000	09:32
Subject Sample	# 2	0.107	09:33
Breath Volume		2.273	09:33
Air Blank		0.000	09:34
Control Test		0.080	09:34
Air Blank		0.000	09:34
Diagnostics Check		OK	09:35

Data Download Date/Time: 01/25/2016 09:55

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EDMOND,COREY,M
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: EDMOND,COREY
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 18:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:39

SUBJECT'S LAST NAME: SANDOVAL SALINAS

DATE/TIME OF TEST: 01/22/2016 20:07:17

Diagnostics Check		OK	20:00
Air Blank		0.000	20:01
Control Test		0.080	20:01
Air Blank		0.000	20:02
Subject Sample	# 1	0.161	20:02
Breath Volume		2.382	20:02
Air Blank		0.000	20:03
Air Blank		0.000	20:05
Subject Sample	# 2	0.152	20:05
Breath Volume		3.601	20:05
Air Blank		0.000	20:06
Control Test		0.080	20:06
Air Blank		0.000	20:07
Diagnostics Check		OK	20:07

Data Download Date/Time: 01/25/2016 09:55

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KENNEDY,LISA,M
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: RALL,T
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 01/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:22

SUBJECT'S LAST NAME: APLARA

DATE/TIME OF TEST: 01/26/2016 23:50:49

Diagnostics Check		OK	23:43
Air Blank		0.000	23:44
Control Test		0.080	23:44
Air Blank		0.000	23:45
Subject Sample	# 1	0.128	23:46
Breath Volume		2.203	23:46
Air Blank		0.000	23:46
Air Blank		0.000	23:48
Subject Sample	# 2	0.121	23:49
Breath Volume		1.882	23:49
Air Blank		0.000	23:49
Control Test		0.079	23:50
Air Blank		0.000	23:50
Diagnostics Check		OK	23:50

Data Download Date/Time: 02/17/2016 08:49

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NAYLOR,MAURICE,E
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: NAYLOR,MAURICE
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07515080A4
DRY GAS EXPIRY DATE: 04/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:06

SUBJECT'S LAST NAME: SANTAMARIA CONTRERAS

DATE/TIME OF TEST: 01/01/2016 00:37:04

Diagnostics Check		OK	00:30
Air Blank		0.000	00:31
Control Test		0.078	00:31
Air Blank		0.000	00:32
Subject Sample	# 1	0.113	00:32
Breath Volume		1.425	00:32
Air Blank		0.000	00:33
Air Blank		0.000	00:35
Subject Sample	# 2	0.102	00:35
Breath Volume		1.656	00:35
Air Blank		0.000	00:36
Control Test		0.079	00:36
Air Blank		0.000	00:36
Diagnostics Check		OK	00:37

Data Download Date/Time: 01/25/2016 12:43

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUNCAN,NICHOLAS,M
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: DUNCAN,NICHOLAS
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07515080A4
DRY GAS EXPIRY DATE: 04/05/2017
VIOLATION CODE: DUI
ARREST TIME: 19:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:40

SUBJECT'S LAST NAME: CONDELLO

DATE/TIME OF TEST: 01/02/2016 21:09:46

Diagnostics Check		OK	21:03
Air Blank		0.000	21:03
Control Test		0.079	21:03
Air Blank		0.000	21:04
Subject Sample	# 1	0.000	21:05
Breath Volume		1.429	21:05
Air Blank		0.000	21:05
Air Blank		0.000	21:07
Subject Sample	# 2	0.000	21:08
Breath Volume		1.847	21:08
Air Blank		0.000	21:08
Control Test		0.077	21:09
Air Blank		0.000	21:09
Diagnostics Check		OK	21:09

Data Download Date/Time: 01/25/2016 12:43

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EDMOND,COREY,M
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: EDMOND,COREY
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07515080A4
DRY GAS EXPIRY DATE: 04/05/2017
VIOLATION CODE: DUI
ARREST TIME: 19:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:16

SUBJECT'S LAST NAME: COONEY

DATE/TIME OF TEST: 01/02/2016 21:46:18

Diagnostics Check		OK	21:40
Air Blank		0.000	21:40
Control Test		0.079	21:40
Air Blank		0.000	21:41
Subject Sample	# 1	0.104	21:41
Breath Volume		3.691	21:41
Air Blank		0.000	21:42
Air Blank		0.000	21:44
Subject Sample	# 2	0.100	21:44
Breath Volume		3.566	21:44
Air Blank		0.000	21:45
Control Test		0.077	21:45
Air Blank		0.000	21:46
Diagnostics Check		OK	21:46

Data Download Date/Time: 01/25/2016 12:43

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUNCAN,NICHOLAS,M
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: DUNCAN,NICHOLAS
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07515080A4
DRY GAS EXPIRY DATE: 04/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: MCDOWELL

DATE/TIME OF TEST: 01/04/2016 02:34:49

Diagnostics Check		OK	02:28
Air Blank		0.000	02:28
Control Test		0.079	02:29
Air Blank		0.000	02:29
Subject Sample	# 1	0.167	02:30
Breath Volume		1.519	02:30
Air Blank		0.000	02:31
Air Blank		0.000	02:32
Subject Sample	# 2	0.180	02:33
Breath Volume		1.945	02:33
Air Blank		0.000	02:33
Control Test		0.078	02:34
Air Blank		0.000	02:34
Diagnostics Check		OK	02:34

Data Download Date/Time: 01/25/2016 12:43

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOMEZ,HENRRY,
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: GOMEZ,HENRRY
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07515080A4
DRY GAS EXPIRY DATE: 04/05/2017
VIOLATION CODE: DUI
ARREST TIME: 17:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:53

SUBJECT'S LAST NAME: PORTER

DATE/TIME OF TEST: 01/04/2016 19:20:24

Diagnostics Check		OK	19:14
Air Blank		0.000	19:14
Control Test		0.079	19:14
Air Blank		0.000	19:15
Subject Sample	# 1	0.068	19:15
Breath Volume		3.730	19:15
Air Blank		0.000	19:16
Air Blank		0.000	19:18
Subject Sample	# 2	0.056	19:18
Breath Volume		1.980	19:18
Air Blank		0.000	19:19
Control Test		0.078	19:19
Air Blank		0.000	19:20
Diagnostics Check		OK	19:20

Data Download Date/Time: 01/25/2016 12:43

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOULD,LEONARD,
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: GOULD,LEONARD
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07515080A4
DRY GAS EXPIRY DATE: 04/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: ISENHOUR

DATE/TIME OF TEST: 01/05/2016 23:46:36

Diagnostics Check		OK	23:40
Air Blank		0.000	23:40
Control Test		0.080	23:41
Air Blank		0.000	23:41
Subject Sample	# 1	0.085	23:42
Breath Volume		1.484	23:42
Air Blank		0.000	23:42
Air Blank		0.000	23:44
Subject Sample	# 2	0.085	23:45
Breath Volume		1.617	23:45
Air Blank		0.000	23:45
Control Test		0.079	23:45
Air Blank		0.000	23:46
Diagnostics Check		OK	23:46

Data Download Date/Time: 01/25/2016 12:43

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRIGGS,CHANCE,R
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: GROSSI,BLAKE
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07515080A4
DRY GAS EXPIRY DATE: 04/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: BRIGGS

DATE/TIME OF TEST: 01/06/2016 01:26:54

Diagnostics Check	OK	01:25	
Air Blank	0.000	01:25	
Control Test	0.078	01:26	
Air Blank		01:26	Improper Sample
Air Blank	0.000	01:26	

Data Download Date/Time: 01/25/2016 12:43

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GROSSI,BLAKE,M
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: GROSSI,BLAKE
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07515080A4
DRY GAS EXPIRY DATE: 04/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: BRIGGS

DATE/TIME OF TEST: 01/06/2016 01:40:32

Diagnostics Check		OK	01:33
Air Blank		0.000	01:34
Control Test		0.079	01:34
Air Blank		0.000	01:35
Subject Sample	# 1	0.167	01:35
Breath Volume		4.070	01:35
Air Blank		0.000	01:36
Air Blank		0.000	01:38
Subject Sample	# 2	0.164	01:38
Breath Volume		3.847	01:38
Air Blank		0.000	01:39
Control Test		0.078	01:39
Air Blank		0.000	01:40
Diagnostics Check		OK	01:40

Data Download Date/Time: 01/25/2016 12:43

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAUGHERTY,KENYATTA,O
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: DAUGHERTY,KENYATTA
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07515080A4
DRY GAS EXPIRY DATE: 04/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:15

SUBJECT'S LAST NAME: HEALEY

DATE/TIME OF TEST: 01/06/2016 21:44:27

Diagnostics Check		OK	21:35
Air Blank		0.000	21:36
Control Test		0.078	21:36
Air Blank		0.000	21:37
Subject Sample	# 1	0.166	21:38
Breath Volume		2.089	21:38
Air Blank		0.000	21:39
Air Blank		0.000	21:40
Subject Sample	# 2	0.183	21:42
Breath Volume		4.164	21:42
Air Blank		0.000	21:43
Control Test		0.078	21:43
Air Blank		0.000	21:44
Diagnostics Check		OK	21:44

Data Download Date/Time: 01/25/2016 12:43

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROMAN,ANTHONY,C
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: JOHNSON,S
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07515080A4
DRY GAS EXPIRY DATE: 04/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:07

SUBJECT'S LAST NAME: FIELDS

DATE/TIME OF TEST: 01/07/2016 03:12:32

Diagnostics Check		OK	03:03
Air Blank		0.000	03:03
Control Test		0.078	03:04
Air Blank		0.000	03:04
Subject Sample	# 1	0.184	03:06
Breath Volume		2.359	03:06
Air Blank		0.000	03:06
Air Blank		0.000	03:08
Subject Sample	# 2	0.171	03:10
Breath Volume		1.425	03:10
Air Blank		0.000	03:11
Control Test		0.077	03:11
Air Blank		0.000	03:12
Diagnostics Check		OK	03:12

Data Download Date/Time: 01/25/2016 12:43

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOULD,LEONARD,
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: GOULD,LEONARD
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07515080A4
DRY GAS EXPIRY DATE: 04/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: SWANSON

DATE/TIME OF TEST: 01/08/2016 23:11:56

Diagnostics Check		OK	23:05
Air Blank		0.000	23:06
Control Test		0.079	23:06
Air Blank		0.000	23:07
Subject Sample	# 1	0.127	23:07
Breath Volume		2.250	23:07
Air Blank		0.000	23:08
Air Blank		0.000	23:09
Subject Sample	# 2	0.129	23:10
Breath Volume		2.378	23:10
Air Blank		0.000	23:11
Control Test		0.079	23:11
Air Blank		0.000	23:11
Diagnostics Check		OK	23:11

Data Download Date/Time: 01/25/2016 12:43

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENEMAN,LESA,
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: BRENEMAN,LESA
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07515080A4
DRY GAS EXPIRY DATE: 04/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:13

SUBJECT'S LAST NAME: GERMANY

DATE/TIME OF TEST: 01/09/2016 01:44:33

Diagnostics Check		OK	01:38
Air Blank		0.000	01:38
Control Test		0.079	01:38
Air Blank		0.000	01:39
Subject Sample	# 1	0.087	01:39
Breath Volume		1.363	01:39
Air Blank		0.000	01:40
Air Blank		0.000	01:42
Subject Sample	# 2	0.091	01:42
Breath Volume		3.394	01:42
Air Blank		0.000	01:43
Control Test		0.079	01:43
Air Blank		0.000	01:44
Diagnostics Check		OK	01:44

Data Download Date/Time: 01/25/2016 12:43

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOULD,LEONARD,
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: GOULD,LEONARD
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07515080A4
DRY GAS EXPIRY DATE: 04/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: GOLDMAN

DATE/TIME OF TEST: 01/09/2016 03:22:42

Diagnostics Check		OK	03:15
Air Blank		0.000	03:16
Control Test		0.079	03:16
Air Blank		0.000	03:17
Subject Sample	# 1	0.120	03:17
Breath Volume		2.503	03:17
Air Blank		0.000	03:18
Air Blank		0.000	03:20
Subject Sample	# 2	0.120	03:21
Breath Volume		2.027	03:21
Air Blank		0.000	03:21
Control Test		0.078	03:22
Air Blank		0.000	03:22
Diagnostics Check		OK	03:22

Data Download Date/Time: 01/25/2016 12:43

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOULD,LEONARD,
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: SUTPHIN,STEVEN
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07515080A4
DRY GAS EXPIRY DATE: 04/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: LARKIN

DATE/TIME OF TEST: 01/09/2016 03:31:39

Diagnostics Check		OK	03:25
Air Blank		0.000	03:25
Control Test		0.078	03:26
Air Blank		0.000	03:26
Subject Sample	# 1	0.142	03:27
Breath Volume		2.140	03:27
Air Blank		0.000	03:27
Air Blank		0.000	03:29
Subject Sample	# 2	0.148	03:30
Breath Volume		1.789	03:30
Air Blank		0.000	03:30
Control Test		0.077	03:31
Air Blank		0.000	03:31
Diagnostics Check		OK	03:31

Data Download Date/Time: 01/25/2016 12:43

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOULD,LEONARD,
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: GOULD,LEONARD
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07515080A4
DRY GAS EXPIRY DATE: 04/05/2017
VIOLATION CODE: DUI
ARREST TIME: 18:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:35

SUBJECT'S LAST NAME: CONRAD

DATE/TIME OF TEST: 01/09/2016 19:04:14

Diagnostics Check		OK	18:58
Air Blank		0.000	18:58
Control Test		0.080	18:59
Air Blank		0.000	18:59
Subject Sample	# 1	0.155	18:59
Breath Volume		1.855	18:59
Air Blank		0.000	19:00
Air Blank		0.000	19:02
Subject Sample	# 2	0.156	19:02
Breath Volume		1.851	19:02
Air Blank		0.000	19:03
Control Test		0.078	19:03
Air Blank		0.000	19:04
Diagnostics Check		OK	19:04

Data Download Date/Time: 01/25/2016 12:43

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NAJARRO,CARLOS,O
BREATH TEST OPERATOR AGENCY: Lee Port Authority
ARRESTING OFFICER: MCGOWAN,PATRICK
ARRESTING OFFICER AGENCY: Lee Port Authority
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07515080A4
DRY GAS EXPIRY DATE: 04/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:14

SUBJECT'S LAST NAME: KAI

DATE/TIME OF TEST: 01/10/2016 00:02:01

Diagnostics Check		OK	23:55
Air Blank		0.000	23:56
Control Test		0.079	23:56
Air Blank		0.000	23:56
Subject Sample	# 2	0.147	23:57
Breath Volume		1.925	23:57
Air Blank		0.000	23:58
Air Blank		0.000	23:59
Subject Sample	# 1	0.136	00:00
Breath Volume		1.703	00:00
Air Blank		0.000	00:01
Control Test		0.078	00:01
Air Blank		0.000	00:01
Diagnostics Check		OK	00:02

Data Download Date/Time: 01/25/2016 12:43

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOULD,LEONARD,
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: GOULD,LEONARD
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07515080A4
DRY GAS EXPIRY DATE: 04/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: PATEL

DATE/TIME OF TEST: 01/10/2016 23:09:14

Diagnostics Check		OK	23:02
Air Blank		0.000	23:03
Control Test		0.078	23:03
Air Blank		0.000	23:04
Subject Sample	# 1	0.200	23:04
Breath Volume		1.742	23:04
Air Blank		0.000	23:05
Air Blank		0.000	23:06
Subject Sample	# 2	0.206	23:07
Breath Volume		2.089	23:07
Air Blank		0.000	23:08
Control Test		0.078	23:08
Air Blank		0.000	23:09
Diagnostics Check		OK	23:09

Data Download Date/Time: 01/25/2016 12:43

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALGER,MICHAEL,C
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: CAPITELLI,C
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07515080A4
DRY GAS EXPIRY DATE: 04/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: DOOLITTLE

DATE/TIME OF TEST: 01/11/2016 01:38:28

Diagnostics Check		OK	01:31
Air Blank		0.000	01:32
Control Test		0.079	01:32
Air Blank		0.000	01:33
Subject Sample	# 1	0.181	01:33
Breath Volume		3.468	01:33
Air Blank		0.000	01:34
Air Blank		0.000	01:36
Subject Sample	# 2	0.181	01:36
Breath Volume		3.746	01:36
Air Blank		0.000	01:37
Control Test		0.077	01:37
Air Blank		0.000	01:38
Diagnostics Check		OK	01:38

Data Download Date/Time: 01/25/2016 12:43

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VENTURI,KRISTYN,L
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: WOLFF,NATHANIEL
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07515080A4
DRY GAS EXPIRY DATE: 04/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:33

SUBJECT'S LAST NAME: PATT

DATE/TIME OF TEST: 01/12/2016 05:16:35

Diagnostics Check		OK	05:10
Air Blank		0.000	05:10
Control Test		0.079	05:11
Air Blank		0.000	05:11
Subject Sample	# 1	0.163	05:12
Breath Volume		3.496	05:12
Air Blank		0.000	05:12
Air Blank		0.000	05:14
Subject Sample	# 2	0.157	05:14
Breath Volume		2.917	05:14
Air Blank		0.000	05:15
Control Test		0.078	05:15
Air Blank		0.000	05:16
Diagnostics Check		OK	05:16

Data Download Date/Time: 01/25/2016 12:43

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUNCAN,NICHOLAS,M
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: DUNCAN,NICHOLAS
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07515080A4
DRY GAS EXPIRY DATE: 04/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:01

SUBJECT'S LAST NAME: BORTOLAMETTI

DATE/TIME OF TEST: 01/13/2016 02:29:36

Diagnostics Check		OK	02:23
Air Blank		0.000	02:23
Control Test		0.079	02:24
Air Blank		0.000	02:24
Subject Sample	# 1	0.120	02:25
Breath Volume		1.839	02:25
Air Blank		0.000	02:25
Air Blank		0.000	02:27
Subject Sample	# 2	0.118	02:27
Breath Volume		1.503	02:27
Air Blank		0.000	02:28
Control Test		0.078	02:28
Air Blank		0.000	02:29
Diagnostics Check		OK	02:29

Data Download Date/Time: 01/25/2016 12:43

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOULD,LEONARD,
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: SUTPHIN,STEPHEN
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07515080A4
DRY GAS EXPIRY DATE: 04/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:38

SUBJECT'S LAST NAME: FERRY

DATE/TIME OF TEST: 01/14/2016 00:07:13

Diagnostics Check		OK	00:01
Air Blank		0.000	00:01
Control Test		0.079	00:01
Air Blank		0.000	00:02
Subject Sample	# 1	0.196	00:02
Breath Volume		2.171	00:02
Air Blank		0.000	00:03
Air Blank		0.000	00:05
Subject Sample	# 2	0.194	00:05
Breath Volume		2.132	00:05
Air Blank		0.000	00:06
Control Test		0.079	00:06
Air Blank		0.000	00:07
Diagnostics Check		OK	00:07

Data Download Date/Time: 01/25/2016 12:43

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,ANDREW,R
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: KITTLESON,MYLES
ARRESTING OFFICER AGENCY: FL Gulf Coast UPD
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07515080A4
DRY GAS EXPIRY DATE: 04/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:44

SUBJECT'S LAST NAME: GROSSBERG-MCNALLY

DATE/TIME OF TEST: 01/14/2016 05:14:44

Diagnostics Check		OK	05:08
Air Blank		0.000	05:08
Control Test		0.078	05:08
Air Blank		0.000	05:09
Subject Sample	# 1	0.174	05:10
Breath Volume		1.441	05:10
Air Blank		0.000	05:10
Air Blank		0.000	05:12
Subject Sample	# 2	0.165	05:13
Breath Volume		1.117	05:13
Air Blank		0.000	05:13
Control Test		0.079	05:14
Air Blank		0.000	05:14
Diagnostics Check		OK	05:14

Data Download Date/Time: 01/25/2016 12:43

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,ANDREW,R
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: SMITH,ANDREW
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07515080A4
DRY GAS EXPIRY DATE: 04/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:52

SUBJECT'S LAST NAME: SCHNITKER

DATE/TIME OF TEST: 01/15/2016 03:24:46

Diagnostics Check		OK	03:18
Air Blank		0.000	03:18
Control Test		0.079	03:18
Air Blank		0.000	03:19
Subject Sample	# 1	0.320	03:19
Breath Volume		3.226	03:19
Air Blank		0.000	03:20
Air Blank		0.000	03:22
Subject Sample	# 2	0.317	03:22
Breath Volume		2.257	03:22
Air Blank		0.000	03:23
Control Test		0.077	03:24
Air Blank		0.000	03:24
Diagnostics Check		OK	03:24

Data Download Date/Time: 01/25/2016 12:43

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROMAN,ANTHONY,C
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: RODRIGUEZ,C
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07515080A4
DRY GAS EXPIRY DATE: 04/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: BLOSE

DATE/TIME OF TEST: 01/17/2016 04:13:16

Diagnostics Check		OK	04:05
Air Blank		0.000	04:06
Control Test		0.079	04:06
Air Blank		0.000	04:07
Subject Sample	# 1	0.163	04:08
Breath Volume		3.511	04:08
Air Blank		0.000	04:08
Air Blank		0.000	04:10
Subject Sample	# 2	0.158	04:11
Breath Volume		2.128	04:11
Air Blank		0.000	04:12
Control Test		0.078	04:12
Air Blank		0.000	04:13
Diagnostics Check		OK	04:13

Data Download Date/Time: 01/25/2016 12:43

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROMAN,ANTHONY,C
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: REYES,G
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07515080A4
DRY GAS EXPIRY DATE: 04/05/2017
VIOLATION CODE: DUI
ARREST TIME: 15:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:12

SUBJECT'S LAST NAME: FREDERIKSEN

DATE/TIME OF TEST: 01/17/2016 16:48:44

Diagnostics Check		OK	16:42
Air Blank		0.000	16:42
Control Test		0.080	16:43
Air Blank		0.000	16:43
Subject Sample	# 1	0.291	16:44
Breath Volume		3.113	16:44
Air Blank		0.000	16:44
Air Blank		0.000	16:46
Subject Sample	# 2	0.286	16:47
Breath Volume		1.687	16:47
Air Blank		0.000	16:47
Control Test		0.078	16:48
Air Blank		0.000	16:48
Diagnostics Check		OK	16:48

Data Download Date/Time: 01/25/2016 12:43

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EDMOND,COREY,M
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: EDMOND,COREY
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07515080A4
DRY GAS EXPIRY DATE: 04/05/2017
VIOLATION CODE: DUI
ARREST TIME: 16:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:01

SUBJECT'S LAST NAME: FINAZZO

DATE/TIME OF TEST: 01/17/2016 19:33:04

Diagnostics Check		OK	19:26
Air Blank		0.000	19:27
Control Test		0.078	19:27
Air Blank		0.000	19:28
Subject Sample	# 1	0.185	19:28
Breath Volume		4.753	19:28
Air Blank		0.000	19:29
Air Blank		0.000	19:31
Subject Sample	# 2	0.181	19:31
Breath Volume		5.082	19:31
Air Blank		0.000	19:32
Control Test		0.077	19:32
Air Blank		0.000	19:32
Diagnostics Check		OK	19:33

Data Download Date/Time: 01/25/2016 12:43

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,ANDREW,R
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: SMITH,ANDREW
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07515080A4
DRY GAS EXPIRY DATE: 04/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: ORTIZ-RIVERA

DATE/TIME OF TEST: 01/22/2016 03:35:32

Diagnostics Check		OK	03:29
Air Blank		0.000	03:29
Control Test		0.078	03:29
Air Blank		0.000	03:30
Subject Sample	# 1	0.222	03:30
Breath Volume		2.921	03:30
Air Blank		0.000	03:31
Air Blank		0.000	03:33
Subject Sample	# 2	0.214	03:33
Breath Volume		3.246	03:33
Air Blank		0.000	03:34
Control Test		0.078	03:34
Air Blank		0.000	03:35
Diagnostics Check		OK	03:35

Data Download Date/Time: 01/25/2016 12:43

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SOMMERS,EDWARD,J
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: SOMMERS,EDWARD
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07515080A4
DRY GAS EXPIRY DATE: 04/05/2017
VIOLATION CODE: DUI
ARREST TIME: 19:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: FLORES-JUEREZ

DATE/TIME OF TEST: 01/23/2016 23:56:55

Diagnostics Check		OK	23:50
Air Blank		0.000	23:51
Control Test		0.078	23:51
Air Blank		0.000	23:52
Subject Sample	# 1	0.163	23:52
Breath Volume		2.203	23:52
Air Blank		0.000	23:53
Air Blank		0.000	23:54
Subject Sample	# 2	0.170	23:55
Breath Volume		3.109	23:55
Air Blank		0.000	23:55
Control Test		0.078	23:56
Air Blank		0.000	23:56
Diagnostics Check		OK	23:56

Data Download Date/Time: 01/25/2016 12:43

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOULD,LEONARD,
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: GOULD,LEONARD
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07515080A4
DRY GAS EXPIRY DATE: 04/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 01/24/2016 01:07:06

Diagnostics Check		OK	01:00
Air Blank		0.000	01:00
Control Test		0.078	01:01
Air Blank		0.000	01:01
Subject Sample	# 1	0.202	01:01
Breath Volume		2.187	01:01
Air Blank		0.000	01:02
Air Blank		0.000	01:04
Subject Sample	# 2	0.182	01:05
Breath Volume		2.675	01:05
Air Blank		0.000	01:06
Control Test		0.078	01:06
Air Blank		0.000	01:07
Diagnostics Check		OK	01:07

Data Download Date/Time: 01/25/2016 12:43

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROMAN,ANTHONY,C
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: MCGOWAN,P
ARRESTING OFFICER AGENCY: Lee Port Authority
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07515080A4
DRY GAS EXPIRY DATE: 04/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: FLORES

DATE/TIME OF TEST: 01/24/2016 23:47:08

Diagnostics Check		OK	23:35	
Air Blank		0.000	23:35	
Control Test		0.079	23:36	
Air Blank		0.000	23:36	
Subject Sample	# 1	0.153	23:39	Volume Not Met
Breath Volume		0.015	23:39	
Air Blank		0.000	23:40	
Air Blank		0.000	23:42	
Subject Sample	# 2	0.141	23:45	Volume Not Met
Breath Volume		0.351	23:45	
Air Blank		0.000	23:46	
Control Test		0.077	23:46	
Air Blank		0.000	23:47	
Diagnostics Check		OK	23:47	

Data Download Date/Time: 01/25/2016 12:43

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROMAN,ANTHONY,C
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: MCGOWAN,P
ARRESTING OFFICER AGENCY: Lee Port Authority
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07515080A4
DRY GAS EXPIRY DATE: 04/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: FLORES

DATE/TIME OF TEST: 01/24/2016 23:56:24

Diagnostics Check		OK	23:49	
Air Blank		0.000	23:49	
Control Test		0.077	23:50	
Air Blank		0.000	23:50	
Subject Sample	# 1	0.130	23:50	
Breath Volume		1.628	23:50	
Air Blank		0.000	23:51	
Air Blank		0.000	23:53	
Subject Sample	# 2		23:54	Subject Test Refused
Breath Volume		0.000	23:54	
Air Blank		0.000	23:55	
Control Test		0.078	23:55	
Air Blank		0.000	23:56	
Diagnostics Check		OK	23:56	

Data Download Date/Time: 01/25/2016 12:43

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUNCAN,NICHOLAS,
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: RALL,T
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 01/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07515080A4
DRY GAS EXPIRY DATE: 04/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: A520864935460

DATE/TIME OF TEST: 01/26/2016 02:44:27

Diagnostics Check		OK	02:38
Air Blank		0.000	02:38
Control Test		0.078	02:38
Air Blank		0.000	02:39
Subject Sample	# 1	0.121	02:39
Breath Volume		2.476	02:39
Air Blank		0.000	02:40
Air Blank		0.000	02:42
Subject Sample	# 2	0.103	02:42
Breath Volume		2.496	02:42
Air Blank		0.000	02:43
Control Test		0.077	02:43
Air Blank		0.000	02:44
Diagnostics Check		OK	02:44

Data Download Date/Time: 02/17/2016 10:46

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUNCAN,NICHOLAS,
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: JOHNSON,S
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 01/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07515080A4
DRY GAS EXPIRY DATE: 04/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:14

SUBJECT'S LAST NAME: NUSBAUM

DATE/TIME OF TEST: 01/26/2016 02:55:10

Diagnostics Check		OK	02:49
Air Blank		0.000	02:49
Control Test		0.078	02:49
Air Blank		0.000	02:50
Subject Sample	# 1	0.067	02:50
Breath Volume		2.175	02:50
Air Blank		0.000	02:51
Air Blank		0.000	02:53
Subject Sample	# 2	0.067	02:53
Breath Volume		2.351	02:53
Air Blank		0.000	02:54
Control Test		0.077	02:54
Air Blank		0.000	02:55
Diagnostics Check		OK	02:55

Data Download Date/Time: 02/17/2016 10:46

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUNCAN,NICHOLAS,
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: DUNCAN,NICHOLAS
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 01/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07515080A4
DRY GAS EXPIRY DATE: 04/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:16

SUBJECT'S LAST NAME: SAILOR

DATE/TIME OF TEST: 01/26/2016 21:46:22

Diagnostics Check		OK	21:40
Air Blank		0.000	21:40
Control Test		0.078	21:40
Air Blank		0.000	21:41
Subject Sample	# 1	0.139	21:41
Breath Volume		1.210	21:41
Air Blank		0.000	21:42
Air Blank		0.000	21:44
Subject Sample	# 2	0.148	21:44
Breath Volume		1.171	21:44
Air Blank		0.000	21:45
Control Test		0.077	21:45
Air Blank		0.000	21:46
Diagnostics Check		OK	21:46

Data Download Date/Time: 02/17/2016 10:46

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CATES,CURTIS,A
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: CATES,CURTIS
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 01/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07515080A4
DRY GAS EXPIRY DATE: 04/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:56

SUBJECT'S LAST NAME: LEONOR-MARTINEZ

DATE/TIME OF TEST: 01/27/2016 22:38:19

Diagnostics Check		OK	22:23	
Air Blank		0.000	22:24	
Control Test		0.079	22:24	
Air Blank		0.000	22:24	
Subject Sample	# 1	0.096	22:28	Volume Not Met
Breath Volume		0.234	22:28	
Air Blank		0.000	22:28	
Air Blank		0.000	22:30	
Subject Sample	# 2	0.097	22:31	
Breath Volume		1.101	22:31	
Air Blank		0.000	22:31	
Air Blank		0.000	22:33	
Subject Sample	# 3		22:36	Subject Test Refused
Breath Volume		0.140	22:36	
Air Blank		0.000	22:37	
Control Test		0.079	22:37	
Air Blank		0.000	22:38	
Diagnostics Check		OK	22:38	

Data Download Date/Time: 02/17/2016 10:46

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CATES,CURTIS,A
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: CATES,CURTIS
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 01/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07515080A4
DRY GAS EXPIRY DATE: 04/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:14

SUBJECT'S LAST NAME: GAGNIER

DATE/TIME OF TEST: 01/28/2016 04:50:13

Diagnostics Check		OK	04:40	
Air Blank		0.000	04:40	
Control Test		0.078	04:41	
Air Blank		0.000	04:41	
Subject Sample	# 1	0.163	04:44	Volume Not Met
Breath Volume		0.105	04:44	
Air Blank		0.000	04:45	
Air Blank		0.000	04:47	
Subject Sample	# 2		04:48	Subject Test Refused
Breath Volume		0.089	04:48	
Air Blank		0.000	04:49	
Control Test		0.076	04:49	
Air Blank		0.000	04:50	
Diagnostics Check		OK	04:50	

Data Download Date/Time: 02/17/2016 10:47

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOULD,LEONARD,
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: GOULD,LEONARD
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 01/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07515080A4
DRY GAS EXPIRY DATE: 04/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: HERRERA

DATE/TIME OF TEST: 01/28/2016 23:57:05

Diagnostics Check		OK	23:51
Air Blank		0.000	23:51
Control Test		0.079	23:51
Air Blank		0.000	23:52
Subject Sample	# 1	0.151	23:52
Breath Volume		1.140	23:52
Air Blank		0.000	23:53
Air Blank		0.000	23:55
Subject Sample	# 2	0.146	23:55
Breath Volume		1.246	23:55
Air Blank		0.000	23:56
Control Test		0.078	23:56
Air Blank		0.000	23:57
Diagnostics Check		OK	23:57

Data Download Date/Time: 02/17/2016 10:47

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VENTURI,KRISTYN,L
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: MANLEY,COREY
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 01/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07515080A4
DRY GAS EXPIRY DATE: 04/05/2017
VIOLATION CODE: DUI
ARREST TIME: 13:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:44

SUBJECT'S LAST NAME: FILLINGAME

DATE/TIME OF TEST: 01/29/2016 14:28:34

Diagnostics Check		OK	14:21
Air Blank		0.000	14:22
Control Test		0.079	14:22
Air Blank		0.000	14:23
Subject Sample	# 1	0.000	14:24
Breath Volume		1.773	14:24
Air Blank		0.000	14:24
Air Blank		0.000	14:26
Subject Sample	# 2	0.000	14:27
Breath Volume		1.808	14:27
Air Blank		0.000	14:27
Control Test		0.078	14:27
Air Blank		0.000	14:28
Diagnostics Check		OK	14:28

Data Download Date/Time: 02/17/2016 10:47

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROMAN,ANTHONY,C
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: REYES,G
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 01/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07515080A4
DRY GAS EXPIRY DATE: 04/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:49

SUBJECT'S LAST NAME: WARGO

DATE/TIME OF TEST: 01/30/2016 23:16:10

Diagnostics Check		OK	23:09
Air Blank		0.000	23:10
Control Test		0.079	23:10
Air Blank		0.000	23:11
Subject Sample	# 1	0.183	23:11
Breath Volume		1.308	23:11
Air Blank		0.000	23:12
Air Blank		0.000	23:14
Subject Sample	# 2	0.184	23:14
Breath Volume		1.914	23:14
Air Blank		0.000	23:15
Control Test		0.078	23:15
Air Blank		0.000	23:16
Diagnostics Check		OK	23:16

Data Download Date/Time: 02/17/2016 10:47

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VERBIEST,RICHARD,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: VERBIEST,RICHARD
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 01/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07515080A4
DRY GAS EXPIRY DATE: 04/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:51

SUBJECT'S LAST NAME: JANOTS

DATE/TIME OF TEST: 01/31/2016 02:23:37

Diagnostics Check		OK	02:14	
Air Blank		0.000	02:14	
Control Test		0.078	02:15	
Air Blank		0.000	02:15	
Subject Sample	# 1		02:18	No Sample Provided
Breath Volume		0.000	02:18	
Air Blank		0.000	02:19	
Air Blank		0.000	02:21	
Subject Sample	# 2		02:22	Subject Test Refused
Breath Volume		0.000	02:22	
Air Blank		0.000	02:22	
Control Test		0.078	02:23	
Air Blank		0.000	02:23	
Diagnostics Check		OK	02:23	

Data Download Date/Time: 02/17/2016 10:47

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VERBIEST,RICHARD,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: VERBIEST,RICHARD
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 01/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07515080A4
DRY GAS EXPIRY DATE: 04/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: JANOTS

DATE/TIME OF TEST: 01/31/2016 03:10:39

Diagnostics Check		OK	02:59	
Air Blank		0.000	02:59	
Control Test		0.078	03:00	
Air Blank		0.000	03:00	
Subject Sample	# 1		03:03	No Sample Provided
Breath Volume		0.000	03:03	
Air Blank		0.000	03:04	
Air Blank		0.000	03:06	
Subject Sample	# 2		03:09	No Sample Provided
Breath Volume		0.000	03:09	
Air Blank		0.000	03:09	
Control Test		0.079	03:10	
Air Blank		0.000	03:10	
Diagnostics Check		OK	03:10	

Data Download Date/Time: 02/17/2016 10:47

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VENTURI,KRISTYN,L
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: ANTOINE,EDWIN
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 01/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07515080A4
DRY GAS EXPIRY DATE: 04/05/2017
VIOLATION CODE: DUI
ARREST TIME: 10:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:10

SUBJECT'S LAST NAME: CASTANEDA CASTANEDA

DATE/TIME OF TEST: 01/31/2016 13:55:00

Diagnostics Check		OK	13:52	
Air Blank		0.000	13:52	
Control Test		0.078	13:52	
Air Blank		0.000	13:53	
Subject Sample	# 1		13:53	Subject Test Refused
Breath Volume		0.000	13:53	
Air Blank		0.000	13:54	
Control Test		0.078	13:54	
Air Blank		0.000	13:54	
Diagnostics Check		OK	13:55	

Data Download Date/Time: 02/17/2016 10:47

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VENTURI,KRISTYN,L
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: VIDAL,JOE
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 01/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07515080A4
DRY GAS EXPIRY DATE: 04/05/2017
VIOLATION CODE: DUI
ARREST TIME: 14:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:40

SUBJECT'S LAST NAME: SILVA

DATE/TIME OF TEST: 01/31/2016 15:18:05

Diagnostics Check		OK	15:10
Air Blank		0.000	15:11
Control Test		0.079	15:11
Air Blank		0.000	15:11
Subject Sample	# 1	0.235	15:13
Breath Volume		1.605	15:13
Air Blank		0.000	15:13
Air Blank		0.000	15:15
Subject Sample	# 2	0.248	15:16
Breath Volume		1.871	15:16
Air Blank		0.000	15:17
Control Test		0.077	15:17
Air Blank		0.000	15:17
Diagnostics Check		OK	15:18

Data Download Date/Time: 02/17/2016 10:47

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROMAN,ANTHOY,C
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: JOHNSON,S
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 01/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07515080A4
DRY GAS EXPIRY DATE: 04/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:56

SUBJECT'S LAST NAME: MCCANN

DATE/TIME OF TEST: 01/31/2016 23:25:53

Diagnostics Check		OK	23:19
Air Blank		0.000	23:19
Control Test		0.078	23:20
Air Blank		0.000	23:20
Subject Sample	# 1	0.179	23:21
Breath Volume		3.722	23:21
Air Blank		0.000	23:22
Air Blank		0.000	23:23
Subject Sample	# 2	0.182	23:24
Breath Volume		3.550	23:24
Air Blank		0.000	23:24
Control Test		0.077	23:25
Air Blank		0.000	23:25
Diagnostics Check		OK	23:25

Data Download Date/Time: 02/17/2016 10:47

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASTILLO,NICKOLAUS,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: LUKASZ,C
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 12/01/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14214080A6
DRY GAS EXPIRY DATE: 07/05/2016
VIOLATION CODE: DUI
ARREST TIME: 01:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:56

SUBJECT'S LAST NAME: FRASER

DATE/TIME OF TEST: 01/01/2016 02:28:00

Diagnostics Check		OK	02:20
Air Blank		0.000	02:21
Control Test		0.078	02:21
Air Blank		0.000	02:22
Subject Sample	# 1	0.145	02:23
Breath Volume		2.121	02:23
Air Blank		0.000	02:23
Air Blank		0.000	02:25
Subject Sample	# 2	0.144	02:26
Breath Volume		1.972	02:26
Air Blank		0.000	02:27
Control Test		0.078	02:27
Air Blank		0.000	02:27
Diagnostics Check		OK	02:28

Data Download Date/Time: 01/04/2016 14:05

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KENNEDY,DANIEL,L
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: KENNEDY,D
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 12/01/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14214080A6
DRY GAS EXPIRY DATE: 07/05/2016
VIOLATION CODE: DUI
ARREST TIME: 16:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:39

SUBJECT'S LAST NAME: WOLLER

DATE/TIME OF TEST: 01/01/2016 19:11:25

Diagnostics Check		OK	18:59	
Air Blank		0.000	19:00	
Control Test		0.078	19:00	
Air Blank		0.000	19:00	
Subject Sample	# 1	0.149	19:04	Volume Not Met
Breath Volume		0.218	19:04	
Air Blank		0.000	19:04	
Air Blank		0.000	19:06	
Subject Sample	# 2	0.155	19:09	Volume Not Met
Breath Volume		0.070	19:09	
Air Blank		0.000	19:10	
Control Test		0.078	19:10	
Air Blank		0.000	19:11	
Diagnostics Check		OK	19:11	

Data Download Date/Time: 01/04/2016 14:05

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHRISTIANSEN,DAVID,M
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: CHRISTIANSEN,DAVID
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 12/01/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14214080A6
DRY GAS EXPIRY DATE: 07/05/2016
VIOLATION CODE: DUI
ARREST TIME: 18:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:22

SUBJECT'S LAST NAME: VASQUES ROBLERO

DATE/TIME OF TEST: 01/01/2016 19:53:56

Diagnostics Check		OK	19:42	
Air Blank		0.000	19:42	
Control Test		0.077	19:43	
Air Blank		0.000	19:43	
Subject Sample	# 1	0.106	19:46	Volume Not Met
Breath Volume		0.015	19:46	
Air Blank		0.000	19:47	
Air Blank		0.000	19:49	
Subject Sample	# 2	0.173	19:52	Volume Not Met
Breath Volume		0.074	19:52	
Air Blank		0.000	19:53	
Control Test		0.077	19:53	
Air Blank		0.000	19:53	
Diagnostics Check		OK	19:53	

Data Download Date/Time: 01/04/2016 14:05

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHRISTIANSEN,DAVID,M
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: CHRISTIANSEN,DAVID
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 12/01/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14214080A6
DRY GAS EXPIRY DATE: 07/05/2016
VIOLATION CODE: DUI
ARREST TIME: 18:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:22

SUBJECT'S LAST NAME: VASQUES ROBLERO

DATE/TIME OF TEST: 01/01/2016 20:04:46

Diagnostics Check		OK	19:56	
Air Blank		0.000	19:56	
Control Test		0.077	19:57	
Air Blank		0.000	19:57	
Subject Sample	# 1	0.185	20:00	Volume Not Met
Breath Volume		0.019	20:00	
Air Blank		0.000	20:01	
Air Blank		0.000	20:03	
Subject Sample	# 2		20:03	Subject Test Refused
Breath Volume		0.000	20:03	
Air Blank		0.000	20:03	
Control Test		0.077	20:04	
Air Blank		0.000	20:04	
Diagnostics Check		OK	20:04	

Data Download Date/Time: 01/04/2016 14:05

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHRISTIANSEN,DAVID,M
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: CHRISTIANSEN,DAVID
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 12/01/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14214080A6
DRY GAS EXPIRY DATE: 07/05/2016
VIOLATION CODE: DUI
ARREST TIME: 03:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:36

SUBJECT'S LAST NAME: GUTIERREZ

DATE/TIME OF TEST: 01/02/2016 05:06:51

Diagnostics Check		OK	05:00
Air Blank		0.000	05:00
Control Test		0.078	05:00
Air Blank		0.000	05:01
Subject Sample	# 1	0.163	05:02
Breath Volume		1.894	05:02
Air Blank		0.000	05:02
Air Blank		0.000	05:04
Subject Sample	# 2	0.162	05:05
Breath Volume		1.273	05:05
Air Blank		0.000	05:05
Control Test		0.077	05:06
Air Blank		0.000	05:06
Diagnostics Check		OK	05:06

Data Download Date/Time: 01/04/2016 14:05

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIBEY,AMANDA,L
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: DELOMBA,GERALD
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 12/01/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14214080A6
DRY GAS EXPIRY DATE: 07/05/2016
VIOLATION CODE: DUI
ARREST TIME: 21:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: SPERA

DATE/TIME OF TEST: 01/02/2016 23:04:33

Diagnostics Check		OK	22:58
Air Blank		0.000	22:58
Control Test		0.079	22:58
Air Blank		0.000	22:59
Subject Sample	# 1	0.084	23:00
Breath Volume		1.781	23:00
Air Blank		0.000	23:00
Air Blank		0.000	23:02
Subject Sample	# 2	0.085	23:03
Breath Volume		1.703	23:03
Air Blank		0.000	23:03
Control Test		0.077	23:03
Air Blank		0.000	23:04
Diagnostics Check		OK	23:04

Data Download Date/Time: 01/04/2016 14:05

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALES,E,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: COLLINS,J
ARRESTING OFFICER AGENCY: Naples PD
LAST AGENCY INSPECTION: 12/01/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14214080A6
DRY GAS EXPIRY DATE: 07/05/2016
VIOLATION CODE: DUI
ARREST TIME: 19:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:33

SUBJECT'S LAST NAME: SAGER

DATE/TIME OF TEST: 01/03/2016 22:03:16

Diagnostics Check		OK	21:56
Air Blank		0.000	21:57
Control Test		0.078	21:57
Air Blank		0.000	21:58
Subject Sample	# 1	0.095	21:58
Breath Volume		4.164	21:58
Air Blank		0.000	21:59
Air Blank		0.000	22:01
Subject Sample	# 2	0.091	22:01
Breath Volume		2.605	22:01
Air Blank		0.000	22:02
Control Test		0.078	22:02
Air Blank		0.000	22:03
Diagnostics Check		OK	22:03

Data Download Date/Time: 01/04/2016 14:05

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOPEZ,P,R
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: DEHAAN,S
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 12/01/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14214080A6
DRY GAS EXPIRY DATE: 07/05/2016
VIOLATION CODE: DUI
ARREST TIME: 22:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:46

SUBJECT'S LAST NAME: SIMMONS

DATE/TIME OF TEST: 01/03/2016 23:22:25

Diagnostics Check		OK	23:12
Air Blank		0.000	23:12
Control Test		0.078	23:13
Air Blank		0.000	23:13
Subject Sample	# 1	0.206	23:14
Breath Volume		1.421	23:14
Air Blank		0.000	23:14
Air Blank		0.000	23:16
Subject Sample	# 2	0.230	23:17
Breath Volume		1.679	23:17
Air Blank		0.000	23:18
Air Blank		0.000	23:20
Subject Sample	# 3	0.219	23:20
Breath Volume		1.371	23:20
Air Blank		0.000	23:21
Control Test		0.077	23:21
Air Blank		0.000	23:22
Diagnostics Check		OK	23:22

Data Download Date/Time: 01/04/2016 14:05

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROGERS,MATTHEW,L
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: DAVENPORT,T
ARRESTING OFFICER AGENCY: Naples PD
LAST AGENCY INSPECTION: 12/01/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14214080A6
DRY GAS EXPIRY DATE: 07/05/2016
VIOLATION CODE: DUI
ARREST TIME: 22:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:33

SUBJECT'S LAST NAME: BRYANT

DATE/TIME OF TEST: 01/04/2016 00:01:50

Diagnostics Check		OK	23:55
Air Blank		0.000	23:55
Control Test		0.077	23:56
Air Blank		0.000	23:56
Subject Sample	# 2	0.118	23:57
Breath Volume		3.250	23:57
Air Blank		0.000	23:57
Air Blank		0.000	23:59
Subject Sample	# 1	0.120	00:00
Breath Volume		3.011	00:00
Air Blank		0.000	00:00
Control Test		0.078	00:01
Air Blank		0.000	00:01
Diagnostics Check		OK	00:01

Data Download Date/Time: 01/04/2016 14:05

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAKER,MATTHEW,R
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: STAUFFER,J
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A4
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:27

SUBJECT'S LAST NAME: CONTRERAS RIVERA

DATE/TIME OF TEST: 01/10/2016 03:58:57

Diagnostics Check		OK	03:51
Air Blank		0.000	03:51
Control Test		0.080	03:52
Air Blank		0.000	03:52
Subject Sample	# 1	0.141	03:53
Breath Volume		1.246	03:53
Air Blank		0.000	03:54
Air Blank		0.000	03:56
Subject Sample	# 2	0.150	03:57
Breath Volume		1.386	03:57
Air Blank		0.000	03:58
Control Test		0.083	03:58
Air Blank		0.000	03:58
Diagnostics Check		OK	03:58

Data Download Date/Time: 02/04/2016 15:40

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAKER,MATTHEW,R
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: MAHOLTZ,S
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A4
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:32

SUBJECT'S LAST NAME: CAPPELEN

DATE/TIME OF TEST: 01/11/2016 01:06:11

Diagnostics Check		OK	00:59
Air Blank		0.000	01:00
Control Test		0.080	01:00
Air Blank		0.000	01:01
Subject Sample	# 1	0.142	01:01
Breath Volume		2.601	01:01
Air Blank		0.000	01:02
Air Blank		0.000	01:04
Subject Sample	# 2	0.139	01:04
Breath Volume		3.472	01:04
Air Blank		0.000	01:05
Control Test		0.080	01:05
Air Blank		0.000	01:06
Diagnostics Check		OK	01:06

Data Download Date/Time: 02/04/2016 15:40

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,KEVIN,J
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: SILL,K
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A4
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:52

SUBJECT'S LAST NAME: SHUMAKER

DATE/TIME OF TEST: 01/11/2016 01:26:34

Diagnostics Check		OK	01:16
Air Blank		0.000	01:17
Control Test		0.079	01:17
Air Blank		0.000	01:18
Subject Sample	# 1	0.145	01:19
Breath Volume		1.253	01:19
Air Blank		0.000	01:20
Air Blank		0.000	01:22
Subject Sample	# 2	0.137	01:24
Breath Volume		1.609	01:24
Air Blank		0.000	01:25
Control Test		0.080	01:25
Air Blank		0.000	01:26
Diagnostics Check		OK	01:26

Data Download Date/Time: 02/04/2016 15:40

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KITTS,R,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: JOHNS,.
ARRESTING OFFICER AGENCY: Naples PD
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 19615080A4
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 16:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:40

SUBJECT'S LAST NAME: STEIN

DATE/TIME OF TEST: 01/11/2016 18:12:13

Diagnostics Check		OK	18:05
Air Blank		0.000	18:05
Control Test		0.080	18:05
Air Blank		0.000	18:06
Subject Sample	# 1	0.000	18:07
Breath Volume		2.816	18:07
Air Blank		0.000	18:08
Air Blank		0.000	18:10
Subject Sample	# 2	0.000	18:10
Breath Volume		2.500	18:10
Air Blank		0.000	18:11
Control Test		0.081	18:11
Air Blank		0.000	18:12
Diagnostics Check		OK	18:12

Data Download Date/Time: 02/04/2016 15:40

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARIMON,KARLHEINZ,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: DOE,JOHN
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A4
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 11:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:40

SUBJECT'S LAST NAME: MEUNINCK

DATE/TIME OF TEST: 01/12/2016 12:15:27

Diagnostics Check		OK	12:08
Air Blank		0.000	12:09
Control Test		0.080	12:09
Air Blank		0.000	12:10
Subject Sample	# 1	0.216	12:10
Breath Volume		1.859	12:10
Air Blank		0.000	12:11
Air Blank		0.000	12:13
Subject Sample	# 2	0.213	12:13
Breath Volume		1.652	12:13
Air Blank		0.000	12:14
Control Test		0.080	12:14
Air Blank		0.000	12:15
Diagnostics Check		OK	12:15

Data Download Date/Time: 02/04/2016 15:40

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FUQUEN,JENNIFER,M
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: ESCOTT,D
ARRESTING OFFICER AGENCY: Naples PD
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A4
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 12:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:20

SUBJECT'S LAST NAME: MOEN

DATE/TIME OF TEST: 01/12/2016 13:53:18

Diagnostics Check		OK	13:46
Air Blank		0.000	13:47
Control Test		0.080	13:47
Air Blank		0.000	13:48
Subject Sample	# 1	0.000	13:48
Breath Volume		4.117	13:48
Air Blank		0.000	13:49
Air Blank		0.000	13:51
Subject Sample	# 2	0.000	13:51
Breath Volume		4.390	13:51
Air Blank		0.000	13:52
Control Test		0.080	13:52
Air Blank		0.000	13:53
Diagnostics Check		OK	13:53

Data Download Date/Time: 02/04/2016 15:40

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEROSA,ANNAMARIE,C
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: DEROSA,ANNAMARIE
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A4
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 18:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:36

SUBJECT'S LAST NAME: PIETRASZEK

DATE/TIME OF TEST: 01/15/2016 20:04:58

Diagnostics Check		OK	19:58
Air Blank		0.000	19:59
Control Test		0.079	19:59
Air Blank		0.000	19:59
Subject Sample	# 1	0.256	20:00
Breath Volume		2.328	20:00
Air Blank		0.000	20:01
Air Blank		0.000	20:03
Subject Sample	# 2	0.269	20:03
Breath Volume		1.972	20:03
Air Blank		0.000	20:04
Control Test		0.079	20:04
Air Blank		0.000	20:04
Diagnostics Check		OK	20:04

Data Download Date/Time: 02/04/2016 15:41

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PATTERSON,D,S
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: MERCADO,D
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A4
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:13

SUBJECT'S LAST NAME: KEMP

DATE/TIME OF TEST: 01/16/2016 02:44:59

Diagnostics Check		OK	02:38
Air Blank		0.000	02:39
Control Test		0.080	02:39
Air Blank		0.000	02:39
Subject Sample	# 1	0.132	02:40
Breath Volume		2.152	02:40
Air Blank		0.000	02:41
Air Blank		0.000	02:43
Subject Sample	# 2	0.129	02:43
Breath Volume		1.500	02:43
Air Blank		0.000	02:44
Control Test		0.080	02:44
Air Blank		0.000	02:44
Diagnostics Check		OK	02:44

Data Download Date/Time: 02/04/2016 15:41

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARRILLO,XAVIER,R
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: SUDANO,J
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A4
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: TOMAS

DATE/TIME OF TEST: 01/16/2016 04:15:03

Diagnostics Check		OK	04:08
Air Blank		0.000	04:08
Control Test		0.080	04:09
Air Blank		0.000	04:09
Subject Sample	# 1	0.121	04:10
Breath Volume		1.406	04:10
Air Blank		0.000	04:10
Air Blank		0.000	04:12
Subject Sample	# 2	0.134	04:13
Breath Volume		1.281	04:13
Air Blank		0.000	04:14
Control Test		0.081	04:14
Air Blank		0.000	04:14
Diagnostics Check		OK	04:15

Data Download Date/Time: 02/04/2016 15:41

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIERMAN,BRENT,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: GROENDYK,R
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A4
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:24

SUBJECT'S LAST NAME: FONTES

DATE/TIME OF TEST: 01/16/2016 04:52:25

Diagnostics Check		OK	04:46
Air Blank		0.000	04:46
Control Test		0.079	04:46
Air Blank		0.000	04:47
Subject Sample	# 1	0.163	04:48
Breath Volume		1.824	04:48
Air Blank		0.000	04:48
Air Blank		0.000	04:50
Subject Sample	# 2	0.164	04:50
Breath Volume		1.242	04:50
Air Blank		0.000	04:51
Control Test		0.080	04:51
Air Blank		0.000	04:52
Diagnostics Check		OK	04:52

Data Download Date/Time: 02/04/2016 15:41

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DARREN,DANIEL,C
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: DARREN,DANIEL
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A4
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: BUI
ARREST TIME: 19:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:20

SUBJECT'S LAST NAME: SCHENK

DATE/TIME OF TEST: 01/16/2016 20:50:01

Diagnostics Check		OK	20:43
Air Blank		0.000	20:43
Control Test		0.080	20:44
Air Blank		0.000	20:44
Subject Sample	# 1	0.041	20:45
Breath Volume		1.480	20:45
Air Blank		0.000	20:46
Air Blank		0.000	20:47
Subject Sample	# 2	0.042	20:48
Breath Volume		2.695	20:48
Air Blank		0.000	20:49
Control Test		0.081	20:49
Air Blank		0.000	20:49
Diagnostics Check		OK	20:50

Data Download Date/Time: 02/04/2016 15:41

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEROSA,ANNAMARIE,C
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: DEROSA,ANNAMRIE
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A4
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:46

SUBJECT'S LAST NAME: ROBB

DATE/TIME OF TEST: 01/17/2016 02:16:34

Diagnostics Check		OK	02:10
Air Blank		0.000	02:10
Control Test		0.080	02:11
Air Blank		0.000	02:11
Subject Sample	# 1	0.110	02:12
Breath Volume		1.320	02:12
Air Blank		0.000	02:12
Air Blank		0.000	02:14
Subject Sample	# 2	0.111	02:15
Breath Volume		1.214	02:15
Air Blank		0.000	02:15
Control Test		0.080	02:15
Air Blank		0.000	02:16
Diagnostics Check		OK	02:16

Data Download Date/Time: 02/04/2016 15:41

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KENNEDY,DANIEL,L
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: KENNEDY,DANIEL
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A4
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:56

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 01/17/2016 23:26:50

Diagnostics Check		OK	23:20
Air Blank		0.000	23:21
Control Test		0.080	23:21
Air Blank		0.000	23:21
Subject Sample	# 1	0.188	23:22
Breath Volume		1.660	23:22
Air Blank		0.000	23:22
Air Blank		0.000	23:24
Subject Sample	# 2	0.185	23:25
Breath Volume		1.804	23:25
Air Blank		0.000	23:25
Control Test		0.080	23:26
Air Blank		0.000	23:26
Diagnostics Check		OK	23:26

Data Download Date/Time: 02/04/2016 15:41

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARBUTO,BLAINE,T
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: TURI,TRACEY
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A4
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: PROB/PAROL
ARREST TIME: 08:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:00

SUBJECT'S LAST NAME: DEAL

DATE/TIME OF TEST: 01/20/2016 08:27:19

Diagnostics Check		OK	08:20
Air Blank		0.000	08:21
Control Test		0.080	08:21
Air Blank		0.000	08:21
Subject Sample	# 1	0.000	08:22
Breath Volume		1.996	08:22
Air Blank		0.000	08:23
Air Blank		0.000	08:25
Subject Sample	# 2	0.000	08:25
Breath Volume		2.007	08:25
Air Blank		0.000	08:26
Control Test		0.081	08:26
Air Blank		0.000	08:27
Diagnostics Check		OK	08:27

Data Download Date/Time: 02/04/2016 15:41

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANGL,B,A
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: REUTHE,R
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A4
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: JIMENEZ-VAZQUEZ

DATE/TIME OF TEST: 01/21/2016 02:33:10

Diagnostics Check		OK	02:26
Air Blank		0.000	02:27
Control Test		0.080	02:27
Air Blank		0.000	02:28
Subject Sample	# 1	0.152	02:28
Breath Volume		2.765	02:28
Air Blank		0.000	02:29
Air Blank		0.000	02:31
Subject Sample	# 2	0.147	02:31
Breath Volume		2.324	02:31
Air Blank		0.000	02:32
Control Test		0.080	02:32
Air Blank		0.000	02:33
Diagnostics Check		OK	02:33

Data Download Date/Time: 02/04/2016 15:41

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROGERS,MATTHEW,L
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: JACKSON,D
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A4
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: DEFLAVIS

DATE/TIME OF TEST: 01/21/2016 04:08:42

Diagnostics Check		OK	03:52	
Air Blank		0.000	03:53	
Control Test		0.080	03:53	
Air Blank		0.000	03:53	
Subject Sample	# 1	0.149	03:57	Volume Not Met
Breath Volume		0.035	03:57	
Air Blank		0.000	03:57	
Air Blank		0.000	03:59	
Subject Sample	# 2	0.147	04:01	
Breath Volume		1.761	04:01	
Air Blank		0.000	04:02	
Air Blank		0.000	04:03	
Subject Sample	# 3	0.130	04:07	Volume Not Met
Breath Volume		0.281	04:07	
Air Blank		0.000	04:07	
Control Test		0.080	04:08	
Air Blank		0.000	04:08	
Diagnostics Check		OK	04:08	

Data Download Date/Time: 02/04/2016 15:41

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROGERS,MATTHEW,L
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: JACKSON,D
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A4
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:11

SUBJECT'S LAST NAME: DEFLAVIS

DATE/TIME OF TEST: 01/21/2016 04:38:46

Diagnostics Check		OK	04:32
Air Blank		0.000	04:32
Control Test		0.079	04:33
Air Blank		0.000	04:33
Subject Sample	# 1	0.136	04:34
Breath Volume		2.144	04:34
Air Blank		0.000	04:34
Air Blank		0.000	04:36
Subject Sample	# 2	0.133	04:37
Breath Volume		2.125	04:37
Air Blank		0.000	04:37
Control Test		0.080	04:38
Air Blank		0.000	04:38
Diagnostics Check		OK	04:38

Data Download Date/Time: 02/04/2016 15:41

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANGL,B,A
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: DEAR,D
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A4
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 19:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:30

SUBJECT'S LAST NAME: CALDERONE

DATE/TIME OF TEST: 01/21/2016 21:04:24

Diagnostics Check		OK	20:58
Air Blank		0.000	20:58
Control Test		0.080	20:59
Air Blank		0.000	20:59
Subject Sample	# 1	0.000	21:00
Breath Volume		2.945	21:00
Air Blank		0.000	21:00
Air Blank		0.000	21:02
Subject Sample	# 2	0.000	21:03
Breath Volume		2.523	21:03
Air Blank		0.000	21:03
Control Test		0.080	21:03
Air Blank		0.000	21:04
Diagnostics Check		OK	21:04

Data Download Date/Time: 02/04/2016 15:41

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BATSON,MITCHELL,J
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: GARCIA,N
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A4
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 18:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:13

SUBJECT'S LAST NAME: QUINN

DATE/TIME OF TEST: 01/22/2016 18:59:39

Diagnostics Check		OK	18:47	
Air Blank		0.000	18:48	
Control Test		0.080	18:48	
Air Blank		0.000	18:49	
Subject Sample	# 1	0.077	18:49	
Breath Volume		2.031	18:49	
Air Blank		0.000	18:50	
Air Blank		0.000	18:51	
Subject Sample	# 2	0.084	18:55	Volume Not Met
Breath Volume		0.945	18:55	
Air Blank		0.000	18:55	
Air Blank		0.000	18:57	
Subject Sample	# 3	0.076	18:58	
Breath Volume		2.480	18:58	
Air Blank		0.000	18:58	
Control Test		0.080	18:59	
Air Blank		0.000	18:59	
Diagnostics Check		OK	18:59	

Data Download Date/Time: 02/04/2016 15:41

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,HANSER,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: STAUFFER,J
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A4
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:02

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 01/23/2016 02:33:46

Diagnostics Check		OK	02:24
Air Blank		0.000	02:25
Control Test		0.079	02:25
Air Blank		0.000	02:25
Subject Sample	# 1	0.218	02:28
Breath Volume		1.382	02:28
Air Blank		0.000	02:29
Air Blank		0.000	02:31
Subject Sample	# 2	0.232	02:32
Breath Volume		1.714	02:32
Air Blank		0.000	02:32
Control Test		0.078	02:33
Air Blank		0.000	02:33
Diagnostics Check		OK	02:33

Data Download Date/Time: 02/04/2016 15:41

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DRISCOLL,JAMES,P
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: DRISCOLL,JAMES
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A4
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:19

SUBJECT'S LAST NAME: FAKRI GRACA

DATE/TIME OF TEST: 01/23/2016 03:45:39

Diagnostics Check		OK	03:39
Air Blank		0.000	03:39
Control Test		0.080	03:39
Air Blank		0.000	03:40
Subject Sample	# 1	0.113	03:41
Breath Volume		2.382	03:41
Air Blank		0.000	03:41
Air Blank		0.000	03:43
Subject Sample	# 2	0.110	03:44
Breath Volume		1.558	03:44
Air Blank		0.000	03:44
Control Test		0.080	03:45
Air Blank		0.000	03:45
Diagnostics Check		OK	03:45

Data Download Date/Time: 02/04/2016 15:41

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DECKER,DAVID,P
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: MCCOY,D
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A4
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: MODUGNO

DATE/TIME OF TEST: 01/24/2016 04:06:56

Diagnostics Check		OK	04:00
Air Blank		0.000	04:00
Control Test		0.080	04:01
Air Blank		0.000	04:01
Subject Sample	# 1	0.235	04:02
Breath Volume		1.539	04:02
Air Blank		0.000	04:03
Air Blank		0.000	04:04
Subject Sample	# 2	0.232	04:05
Breath Volume		1.976	04:05
Air Blank		0.000	04:06
Control Test		0.080	04:06
Air Blank		0.000	04:06
Diagnostics Check		OK	04:06

Data Download Date/Time: 02/04/2016 15:41

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARMENT,KERRY,A
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: DESIMONE,*
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A4
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: OCONNOR

DATE/TIME OF TEST: 01/24/2016 22:51:37

Diagnostics Check		OK	22:44
Air Blank		0.000	22:45
Control Test		0.079	22:45
Air Blank		0.000	22:46
Subject Sample	# 1	0.168	22:46
Breath Volume		1.519	22:46
Air Blank		0.000	22:47
Air Blank		0.000	22:49
Subject Sample	# 2	0.170	22:49
Breath Volume		1.511	22:49
Air Blank		0.000	22:50
Control Test		0.079	22:51
Air Blank		0.000	22:51
Diagnostics Check		OK	22:51

Data Download Date/Time: 02/04/2016 15:41

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COMBS,SEAN,F
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: COMBS,SEAN
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A4
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: OILER

DATE/TIME OF TEST: 01/25/2016 02:02:21

Diagnostics Check		OK	01:55
Air Blank		0.000	01:55
Control Test		0.081	01:56
Air Blank		0.000	01:56
Subject Sample	# 1	0.066	01:57
Breath Volume		1.996	01:57
Air Blank		0.000	01:58
Air Blank		0.000	02:00
Subject Sample	# 2	0.062	02:00
Breath Volume		1.683	02:00
Air Blank		0.000	02:01
Control Test		0.081	02:01
Air Blank		0.000	02:02
Diagnostics Check		OK	02:02

Data Download Date/Time: 02/04/2016 15:41

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELMORE,DALE,A
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: ELMORE,D
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A4
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 12:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:00

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 01/27/2016 13:28:26

Diagnostics Check		OK	13:22
Air Blank		0.000	13:22
Control Test		0.078	13:23
Air Blank		0.000	13:23
Subject Sample	# 1	0.000	13:24
Breath Volume		2.542	13:24
Air Blank		0.000	13:24
Air Blank		0.000	13:26
Subject Sample	# 2	0.000	13:27
Breath Volume		3.242	13:27
Air Blank		0.000	13:27
Control Test		0.080	13:27
Air Blank		0.000	13:28
Diagnostics Check		OK	13:28

Data Download Date/Time: 02/04/2016 15:41

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DRISCOLL,JAMES,P
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: DRISCOLL,JAMES
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A4
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: BONACHEA

DATE/TIME OF TEST: 01/28/2016 01:12:38

Diagnostics Check		OK	01:06
Air Blank		0.000	01:06
Control Test		0.079	01:06
Air Blank		0.000	01:07
Subject Sample	# 1	0.188	01:07
Breath Volume		1.617	01:07
Air Blank		0.000	01:08
Air Blank		0.000	01:10
Subject Sample	# 2	0.208	01:11
Breath Volume		2.101	01:11
Air Blank		0.000	01:11
Control Test		0.080	01:12
Air Blank		0.000	01:12
Diagnostics Check		OK	01:12

Data Download Date/Time: 02/04/2016 15:41

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORNELIUS,TIM,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: CORNELIUS,TIM
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A4
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:52

SUBJECT'S LAST NAME: FEIGHT

DATE/TIME OF TEST: 01/28/2016 22:23:21

Diagnostics Check		OK	22:16
Air Blank		0.000	22:17
Control Test		0.080	22:17
Air Blank		0.000	22:17
Subject Sample	# 1	0.102	22:18
Breath Volume		2.718	22:18
Air Blank		0.000	22:19
Air Blank		0.000	22:21
Subject Sample	# 2	0.101	22:21
Breath Volume		3.324	22:21
Air Blank		0.000	22:22
Control Test		0.081	22:22
Air Blank		0.000	22:23
Diagnostics Check		OK	22:23

Data Download Date/Time: 02/04/2016 15:41

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PATTERSON,D,S
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: MERCADO,D
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A4
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: KURETH

DATE/TIME OF TEST: 01/30/2016 02:08:12

Diagnostics Check		OK	02:01
Air Blank		0.000	02:02
Control Test		0.080	02:02
Air Blank		0.000	02:03
Subject Sample	# 1	0.139	02:03
Breath Volume		1.652	02:03
Air Blank		0.000	02:04
Air Blank		0.000	02:06
Subject Sample	# 2	0.148	02:06
Breath Volume		1.746	02:06
Air Blank		0.000	02:07
Control Test		0.081	02:07
Air Blank		0.000	02:08
Diagnostics Check		OK	02:08

Data Download Date/Time: 02/04/2016 15:41

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOGAN,TIFFANY,J
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: KLEIS,K
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 19615080A4
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 19:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:56

SUBJECT'S LAST NAME: DOMINGUEZ

DATE/TIME OF TEST: 01/30/2016 21:29:39

Diagnostics Check		OK	21:21
Air Blank		0.000	21:21
Control Test		0.080	21:22
Air Blank		0.000	21:22
Subject Sample	# 1	0.098	21:25
Breath Volume		2.765	21:25
Air Blank		0.000	21:25
Air Blank		0.000	21:27
Subject Sample	# 2	0.099	21:28
Breath Volume		1.492	21:28
Air Blank		0.000	21:28
Control Test		0.079	21:29
Air Blank		0.000	21:29
Diagnostics Check		OK	21:29

Data Download Date/Time: 02/04/2016 15:41

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DICARLO,J,T
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: NOEL,M
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A4
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:39

SUBJECT'S LAST NAME: COSTA

DATE/TIME OF TEST: 01/31/2016 00:12:56

Diagnostics Check		OK	00:04
Air Blank		0.000	00:05
Control Test		0.081	00:05
Air Blank		0.000	00:06
Subject Sample	# 1	0.093	00:07
Breath Volume		2.406	00:07
Air Blank		0.000	00:08
Air Blank		0.000	00:10
Subject Sample	# 2	0.080	00:11
Breath Volume		2.554	00:11
Air Blank		0.000	00:12
Control Test		0.079	00:12
Air Blank		0.000	00:12
Diagnostics Check		OK	00:12

Data Download Date/Time: 02/04/2016 15:41

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIERMAN,BRENT,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: BEAIRD,C
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A4
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:13

SUBJECT'S LAST NAME: BERRY

DATE/TIME OF TEST: 01/31/2016 00:40:19

Diagnostics Check		OK	00:33
Air Blank		0.000	00:34
Control Test		0.079	00:34
Air Blank		0.000	00:34
Subject Sample	# 1	0.204	00:35
Breath Volume		2.417	00:35
Air Blank		0.000	00:36
Air Blank		0.000	00:38
Subject Sample	# 2	0.209	00:38
Breath Volume		1.933	00:38
Air Blank		0.000	00:39
Control Test		0.078	00:39
Air Blank		0.000	00:40
Diagnostics Check		OK	00:40

Data Download Date/Time: 02/04/2016 15:41

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARRILLO,XAVIER,R
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: REUTHE,R
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A4
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:47

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 01/31/2016 03:13:16

Diagnostics Check		OK	03:07
Air Blank		0.000	03:07
Control Test		0.080	03:07
Air Blank		0.000	03:08
Subject Sample	# 1	0.146	03:08
Breath Volume		1.882	03:08
Air Blank		0.000	03:09
Air Blank		0.000	03:11
Subject Sample	# 2	0.136	03:11
Breath Volume		2.023	03:11
Air Blank		0.000	03:12
Control Test		0.081	03:12
Air Blank		0.000	03:13
Diagnostics Check		OK	03:13

Data Download Date/Time: 02/04/2016 15:41

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIAZ,DEPUTY,M
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: RODRIGUEZ,R
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 12/01/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14214080A6
DRY GAS EXPIRY DATE: 07/05/2016
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:06

SUBJECT'S LAST NAME: ANDRES ANTONIO

DATE/TIME OF TEST: 01/02/2016 03:34:07

Diagnostics Check		OK	03:28
Air Blank		0.000	03:28
Control Test		0.078	03:28
Air Blank		0.000	03:29
Subject Sample	# 1	0.230	03:29
Breath Volume		2.484	03:29
Air Blank		0.000	03:30
Air Blank		0.000	03:32
Subject Sample	# 2	0.228	03:32
Breath Volume		2.597	03:32
Air Blank		0.000	03:33
Control Test		0.078	03:33
Air Blank		0.000	03:34
Diagnostics Check		OK	03:34

Data Download Date/Time: 01/04/2016 17:09

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCALL,LACEY,J
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: RODRIGUEZ,ROLDAN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14214080A6
DRY GAS EXPIRY DATE: 07/05/2016
VIOLATION CODE: DUI
ARREST TIME: 02:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:46

SUBJECT'S LAST NAME: RIVERA ROMAN

DATE/TIME OF TEST: 01/09/2016 03:13:01

Diagnostics Check		OK	03:07	
Air Blank		0.000	03:08	
Control Test		0.079	03:08	
Air Blank		0.000	03:09	
Subject Sample	# 1		03:11	Subject Test Refused
Breath Volume		0.414	03:11	
Air Blank		0.000	03:12	
Control Test		0.079	03:12	
Air Blank		0.000	03:12	
Diagnostics Check		OK	03:13	

Data Download Date/Time: 02/08/2016 11:09

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SOTO,SAMUEL,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: WILSON,ANTHONY
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 01/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14214080A6
DRY GAS EXPIRY DATE: 07/05/2016
VIOLATION CODE: DUI
ARREST TIME: 00:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:38

SUBJECT'S LAST NAME: MENDOZA

DATE/TIME OF TEST: 01/14/2016 01:05:57

Diagnostics Check		OK	00:59
Air Blank		0.000	00:59
Control Test		0.080	01:00
Air Blank		0.000	01:00
Subject Sample	# 1	0.169	01:01
Breath Volume		1.402	01:01
Air Blank		0.000	01:02
Air Blank		0.000	01:04
Subject Sample	# 2	0.184	01:04
Breath Volume		1.480	01:04
Air Blank		0.000	01:05
Control Test		0.079	01:05
Air Blank		0.000	01:05
Diagnostics Check		OK	01:05

Data Download Date/Time: 02/08/2016 11:09

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000945

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FISHER,SCOTT,R
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: WRASSE,DEPUTY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 01/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:48

SUBJECT'S LAST NAME: LONG

DATE/TIME OF TEST: 01/30/2016 03:24:10

Diagnostics Check		OK	03:10	
Air Blank		0.000	03:10	
Control Test		0.082	03:10	
Air Blank		0.000	03:11	
Subject Sample	# 1		03:14	No Sample Provided
Breath Volume		0.000	03:14	
Air Blank		0.000	03:14	
Air Blank		0.000	03:16	
Subject Sample	# 2	0.270	03:18	
Breath Volume		1.890	03:18	
Air Blank		0.000	03:19	
Air Blank		0.000	03:21	
Subject Sample	# 3	0.264	03:22	
Breath Volume		1.601	03:22	
Air Blank		0.000	03:23	
Control Test		0.082	03:23	
Air Blank		0.000	03:24	
Diagnostics Check		OK	03:24	

Data Download Date/Time: 02/25/2016 19:51

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000945

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARRELL,THOMAS,R
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: NOVA,DEPUTY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 01/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:12

SUBJECT'S LAST NAME: BURNHAM

DATE/TIME OF TEST: 01/30/2016 03:46:12

Diagnostics Check		OK	03:38	
Air Blank		0.000	03:38	
Control Test		0.080	03:38	
Air Blank		0.000	03:39	
Subject Sample	# 1		03:42	No Sample Provided
Breath Volume		0.000	03:42	
Air Blank		0.000	03:42	
Air Blank			03:44	RFI Detect
Air Blank		0.000	03:45	
Control Test		0.081	03:45	
Air Blank		0.000	03:46	
Diagnostics Check		OK	03:46	

Data Download Date/Time: 02/25/2016 19:51

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000945

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FISHER,SCOTT,R
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: EDWARDS,NATE
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 01/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:38

SUBJECT'S LAST NAME: MCCREERY KEMP

DATE/TIME OF TEST: 01/30/2016 04:06:39

Diagnostics Check		OK	03:59
Air Blank		0.000	03:59
Control Test		0.082	03:59
Air Blank		0.000	04:00
Subject Sample	# 1	0.206	04:01
Breath Volume		1.929	04:01
Air Blank		0.000	04:01
Air Blank		0.000	04:03
Subject Sample	# 2	0.201	04:05
Breath Volume		1.695	04:05
Air Blank		0.000	04:05
Control Test		0.081	04:06
Air Blank		0.000	04:06
Diagnostics Check		OK	04:06

Data Download Date/Time: 02/25/2016 19:51

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000946

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHAPMAN,VANESSA,N
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: PATTEN,JACOB
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 12/29/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:13

SUBJECT'S LAST NAME: SHACKELFORD

DATE/TIME OF TEST: 01/01/2016 02:44:08

Diagnostics Check		OK	02:35
Air Blank		0.000	02:36
Control Test		0.080	02:36
Air Blank		0.000	02:37
Subject Sample	# 1	0.103	02:39
Breath Volume		1.390	02:39
Air Blank		0.000	02:40
Air Blank		0.000	02:42
Subject Sample	# 2	0.104	02:42
Breath Volume		1.210	02:42
Air Blank		0.000	02:43
Control Test		0.079	02:43
Air Blank		0.000	02:44
Diagnostics Check		OK	02:44

Data Download Date/Time: 01/28/2016 20:46

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000946

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FISHER,SCOTT,R
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: SELLA,DEPUTY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 12/29/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:06

SUBJECT'S LAST NAME: RAJALA

DATE/TIME OF TEST: 01/02/2016 00:39:01

Diagnostics Check		OK	00:27	
Air Blank		0.000	00:28	
Control Test		0.078	00:28	
Air Blank		0.000	00:28	
Subject Sample	# 1		00:32	No Sample Provided
Breath Volume		0.000	00:32	
Air Blank		0.000	00:32	
Air Blank		0.000	00:34	
Subject Sample	# 2		00:37	No Sample Provided
Breath Volume		0.000	00:37	
Air Blank		0.000	00:38	
Control Test		0.080	00:38	
Air Blank		0.000	00:38	
Diagnostics Check		OK	00:39	

Data Download Date/Time: 01/28/2016 20:46

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000946

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARRELL,THOMAS,R
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: RENZ,CPL
ARRESTING OFFICER AGENCY: Punta Gorda PD
LAST AGENCY INSPECTION: 12/29/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:39

SUBJECT'S LAST NAME: PIERRE

DATE/TIME OF TEST: 01/02/2016 04:25:55

Diagnostics Check		OK	04:13	
Air Blank		0.000	04:14	
Control Test		0.080	04:14	
Air Blank		0.000	04:14	
Subject Sample	# 1	0.086	04:18	Volume Not Met
Breath Volume		0.921	04:18	
Air Blank		0.000	04:18	
Air Blank		0.000	04:20	
Subject Sample	# 2	0.087	04:21	
Breath Volume		1.429	04:21	
Air Blank		0.000	04:21	
Air Blank		0.000	04:23	
Subject Sample	# 3	0.086	04:24	
Breath Volume		1.148	04:24	
Air Blank		0.000	04:25	
Control Test		0.078	04:25	
Air Blank		0.000	04:25	
Diagnostics Check		OK	04:25	

Data Download Date/Time: 01/28/2016 20:46

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000946

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FISHER,SCOTT,R
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: RENZ,OFFICER
ARRESTING OFFICER AGENCY: Punta Gorda PD
LAST AGENCY INSPECTION: 12/29/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:30

SUBJECT'S LAST NAME: TANKSLEY

DATE/TIME OF TEST: 01/02/2016 21:06:01

Diagnostics Check		OK	20:52	
Air Blank		0.000	20:53	
Control Test		0.079	20:53	
Air Blank		0.000	20:53	
Subject Sample	# 1	0.179	20:57	Volume Not Met
Breath Volume		0.000	20:57	
Air Blank		0.000	20:57	
Air Blank		0.000	20:59	
Subject Sample	# 2	0.244	21:00	
Breath Volume		3.210	21:00	
Air Blank		0.000	21:01	
Air Blank		0.000	21:03	
Subject Sample	# 3	0.234	21:04	
Breath Volume		2.519	21:04	
Air Blank		0.000	21:05	
Control Test		0.078	21:05	
Air Blank		0.000	21:05	
Diagnostics Check		OK	21:06	

Data Download Date/Time: 01/28/2016 20:46

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000946

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARRELL,THOMAS,R
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: SELLA,DEPUTY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 12/29/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:58

SUBJECT'S LAST NAME: MURPHY

DATE/TIME OF TEST: 01/02/2016 22:31:03

Diagnostics Check		OK	22:21
Air Blank		0.000	22:21
Control Test		0.079	22:21
Air Blank		0.000	22:22
Subject Sample	# 1	0.265	22:22
Breath Volume		3.167	22:22
Air Blank		0.000	22:23
Air Blank		0.000	22:25
Subject Sample	# 2	0.236	22:26
Breath Volume		1.398	22:26
Air Blank		0.000	22:27
Air Blank		0.000	22:28
Subject Sample	# 3	0.239	22:29
Breath Volume		1.328	22:29
Air Blank		0.000	22:30
Control Test		0.079	22:30
Air Blank		0.000	22:30
Diagnostics Check		OK	22:31

Data Download Date/Time: 01/28/2016 20:46

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000946

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FISHER,SCOTT,R
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: MARTIN,DEPUTY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 12/29/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:07

SUBJECT'S LAST NAME: PARKINSON

DATE/TIME OF TEST: 01/03/2016 01:39:00

Diagnostics Check		OK	01:27	
Air Blank		0.000	01:28	
Control Test		0.080	01:28	
Air Blank		0.000	01:28	
Subject Sample	# 1	0.000	01:31	Volume Not Met
Breath Volume		0.015	01:31	
Air Blank		0.000	01:32	
Air Blank		0.000	01:34	
Subject Sample	# 2	0.000	01:37	Volume Not Met
Breath Volume		0.101	01:37	
Air Blank		0.000	01:38	
Control Test		0.079	01:38	
Air Blank		0.000	01:38	
Diagnostics Check		OK	01:39	

Data Download Date/Time: 01/28/2016 20:46

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000946

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARRELL,THOMAS,R
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: MARTIN,TROOPER
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 12/29/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: LANGE

DATE/TIME OF TEST: 01/03/2016 02:06:46

Diagnostics Check		OK	01:55	
Air Blank		0.000	01:55	
Control Test		0.078	01:56	
Air Blank		0.000	01:56	
Subject Sample	# 1	0.230	01:59	Volume Not Met
Breath Volume		0.015	01:59	
Air Blank		0.000	02:00	
Air Blank		0.000	02:02	
Subject Sample	# 2	0.221	02:05	Volume Not Met
Breath Volume		0.039	02:05	
Air Blank		0.000	02:05	
Control Test		0.078	02:06	
Air Blank		0.000	02:06	
Diagnostics Check		OK	02:06	

Data Download Date/Time: 01/28/2016 20:46

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000946

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EGAN,MELISSA,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: JONES,J
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 12/29/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:47

SUBJECT'S LAST NAME: ROLLERI

DATE/TIME OF TEST: 01/07/2016 01:14:32

Diagnostics Check		OK	01:07
Air Blank		0.000	01:08
Control Test		0.079	01:08
Air Blank		0.000	01:09
Subject Sample	# 1	0.183	01:09
Breath Volume		2.128	01:09
Air Blank		0.000	01:10
Air Blank		0.000	01:12
Subject Sample	# 2	0.190	01:13
Breath Volume		2.519	01:13
Air Blank		0.000	01:13
Control Test		0.079	01:13
Air Blank		0.000	01:14
Diagnostics Check		OK	01:14

Data Download Date/Time: 01/28/2016 20:46

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000946

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EGAN,MELISSA,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: MCDONALD,J
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 12/29/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:42

SUBJECT'S LAST NAME: SNYDER

DATE/TIME OF TEST: 01/08/2016 21:10:05

Diagnostics Check		OK	21:03
Air Blank		0.000	21:03
Control Test		0.080	21:04
Air Blank		0.000	21:04
Subject Sample	# 1	0.273	21:05
Breath Volume		3.515	21:05
Air Blank		0.000	21:06
Air Blank		0.000	21:07
Subject Sample	# 2	0.254	21:08
Breath Volume		2.917	21:08
Air Blank		0.000	21:09
Control Test		0.079	21:09
Air Blank		0.000	21:10
Diagnostics Check		OK	21:10

Data Download Date/Time: 01/28/2016 20:46

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000946

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EGAN,MELISSA,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: MCDONALD,J
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 12/29/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: GRAHAM-SCHROEDER

DATE/TIME OF TEST: 01/10/2016 03:54:56

Diagnostics Check		OK	03:47
Air Blank		0.000	03:48
Control Test		0.080	03:48
Air Blank		0.000	03:48
Subject Sample	# 1	0.236	03:49
Breath Volume		1.660	03:49
Air Blank		0.000	03:50
Air Blank		0.000	03:51
Subject Sample	# 2	0.225	03:53
Breath Volume		2.062	03:53
Air Blank		0.000	03:54
Control Test		0.080	03:54
Air Blank		0.000	03:54
Diagnostics Check		OK	03:54

Data Download Date/Time: 01/28/2016 20:46

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000946

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARRELL,THOMAS,R
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: NOVA,DEPUTY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 12/29/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 01/11/2016 23:26:44

Diagnostics Check		OK	23:20
Air Blank		0.000	23:20
Control Test		0.079	23:21
Air Blank		0.000	23:21
Subject Sample	# 1	0.136	23:22
Breath Volume		2.406	23:22
Air Blank		0.000	23:22
Air Blank		0.000	23:24
Subject Sample	# 2	0.134	23:25
Breath Volume		2.878	23:25
Air Blank		0.000	23:25
Control Test		0.080	23:26
Air Blank		0.000	23:26
Diagnostics Check		OK	23:26

Data Download Date/Time: 01/28/2016 20:46

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000946

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARRELL,THOMAS,R
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: EDWARDS,DEPUTY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 12/29/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:12

SUBJECT'S LAST NAME: AGAN

DATE/TIME OF TEST: 01/12/2016 02:39:04

Diagnostics Check		OK	02:32
Air Blank		0.000	02:33
Control Test		0.080	02:33
Air Blank		0.000	02:33
Subject Sample	# 1	0.200	02:34
Breath Volume		4.101	02:34
Air Blank		0.000	02:34
Air Blank		0.000	02:36
Subject Sample	# 2	0.200	02:37
Breath Volume		3.648	02:37
Air Blank		0.000	02:38
Control Test		0.079	02:38
Air Blank		0.000	02:38
Diagnostics Check		OK	02:39

Data Download Date/Time: 01/28/2016 20:46

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000946

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARRELL,THOMAS,R
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: PUTMAN,DEPUTY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 12/29/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:28

SUBJECT'S LAST NAME: HAKES

DATE/TIME OF TEST: 01/16/2016 00:54:46

Diagnostics Check		OK	00:48
Air Blank		0.000	00:49
Control Test		0.079	00:49
Air Blank		0.000	00:49
Subject Sample	# 1	0.048	00:50
Breath Volume		3.648	00:50
Air Blank		0.000	00:50
Air Blank		0.000	00:52
Subject Sample	# 2	0.050	00:53
Breath Volume		3.886	00:53
Air Blank		0.000	00:53
Control Test		0.080	00:54
Air Blank		0.000	00:54
Diagnostics Check		OK	00:54

Data Download Date/Time: 01/28/2016 20:46

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000946

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARIAS,MONICA,S
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: BALTIMORE,DEPUTY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 12/29/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 09:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:28

SUBJECT'S LAST NAME: BREFKA

DATE/TIME OF TEST: 01/16/2016 10:57:28

Diagnostics Check		OK	10:51
Air Blank		0.000	10:52
Control Test		0.079	10:52
Air Blank		0.000	10:52
Subject Sample	# 1	0.000	10:53
Breath Volume		1.804	10:53
Air Blank		0.000	10:53
Air Blank		0.000	10:55
Subject Sample	# 2	0.000	10:56
Breath Volume		1.820	10:56
Air Blank		0.000	10:56
Control Test		0.079	10:56
Air Blank		0.000	10:57
Diagnostics Check		OK	10:57

Data Download Date/Time: 01/28/2016 20:46

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000946

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EGAN,MELISSA,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: ALLEN,M
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 12/29/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:34

SUBJECT'S LAST NAME: GJEDREM

DATE/TIME OF TEST: 01/19/2016 02:01:33

Diagnostics Check		OK	01:55
Air Blank		0.000	01:56
Control Test		0.079	01:56
Air Blank		0.000	01:56
Subject Sample	# 1	0.047	01:57
Breath Volume		2.152	01:57
Air Blank		0.000	01:57
Air Blank		0.000	01:59
Subject Sample	# 2	0.047	02:00
Breath Volume		1.753	02:00
Air Blank		0.000	02:00
Control Test		0.080	02:00
Air Blank		0.000	02:01
Diagnostics Check		OK	02:01

Data Download Date/Time: 01/28/2016 20:46

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000946

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EGAN,MELISSA,
BREATH TEST OPERATOR AGENCY: Punta Gorda PD
ARRESTING OFFICER: KLEIVER,J
ARRESTING OFFICER AGENCY: Punta Gorda PD
LAST AGENCY INSPECTION: 12/29/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:21

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 01/19/2016 21:49:52

Diagnostics Check		OK	21:43
Air Blank		0.000	21:43
Control Test		0.080	21:43
Air Blank		0.000	21:44
Subject Sample	# 1	0.127	21:45
Breath Volume		1.621	21:45
Air Blank		0.000	21:45
Air Blank		0.000	21:47
Subject Sample	# 2	0.133	21:48
Breath Volume		1.613	21:48
Air Blank		0.000	21:48
Control Test		0.080	21:49
Air Blank		0.000	21:49
Diagnostics Check		OK	21:49

Data Download Date/Time: 01/28/2016 20:46

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000946

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHAPMAN,VANESSA,N
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: MCDONALD,JUSTIN
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 12/29/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:13

SUBJECT'S LAST NAME: CIGICH

DATE/TIME OF TEST: 01/23/2016 04:44:05

Diagnostics Check		OK	04:38
Air Blank		0.000	04:38
Control Test		0.079	04:38
Air Blank		0.000	04:39
Subject Sample	# 1	0.217	04:39
Breath Volume		2.699	04:39
Air Blank		0.000	04:40
Air Blank		0.000	04:42
Subject Sample	# 2	0.234	04:42
Breath Volume		3.589	04:42
Air Blank		0.000	04:43
Control Test		0.079	04:43
Air Blank		0.000	04:44
Diagnostics Check		OK	04:44

Data Download Date/Time: 01/28/2016 20:46

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000946

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FISHER,SCOTT,R
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: SELLA,DEPUTY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 12/29/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 19:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:45

SUBJECT'S LAST NAME: LEWIS

DATE/TIME OF TEST: 01/25/2016 21:12:16

Diagnostics Check		OK	21:05	
Air Blank		0.000	21:06	
Control Test		0.079	21:06	
Air Blank		0.000	21:07	
Subject Sample	# 1	0.275	21:07	
Breath Volume		3.800	21:07	
Air Blank		0.000	21:08	
Air Blank		0.000	21:10	
Subject Sample	# 2	0.269	21:10	
Breath Volume		4.062	21:10	
Air Blank		0.000	21:11	
Control Test		0.079	21:11	
Air Blank			21:11	RFI Detect
Air Blank		0.000	21:12	

Data Download Date/Time: 01/28/2016 20:46

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000947

INSTRUMENT REGISTERED TO: GLADES COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OKUPSKI,RYAN,T
BREATH TEST OPERATOR AGENCY: Glades County SO
ARRESTING OFFICER: OKUPSKI,R
ARRESTING OFFICER AGENCY: Glades County SO
LAST AGENCY INSPECTION: 12/31/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13314080A3
DRY GAS EXPIRY DATE: 06/05/2016
VIOLATION CODE: DUI
ARREST TIME: 15:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:00

SUBJECT'S LAST NAME: ALLEN

DATE/TIME OF TEST: 01/11/2016 16:36:35

Diagnostics Check		OK	16:33	
Air Blank		0.000	16:34	
Control Test		0.077	16:34	
Air Blank		0.000	16:35	
Subject Sample	# 1		16:35	Subject Test Refused
Breath Volume		0.000	16:35	
Air Blank		0.000	16:35	
Control Test		0.077	16:35	
Air Blank		0.000	16:36	
Diagnostics Check		OK	16:36	

Data Download Date/Time: 01/31/2016 10:48

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000947

INSTRUMENT REGISTERED TO: GLADES COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCQUAID,RITA,S
BREATH TEST OPERATOR AGENCY: Glades County SO
ARRESTING OFFICER: MCQUAID,DEWAYNE
ARRESTING OFFICER AGENCY: Glades County SO
LAST AGENCY INSPECTION: 12/31/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13314080A3
DRY GAS EXPIRY DATE: 06/05/2016
VIOLATION CODE: ADMIN
ARREST TIME: 19:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: ALLEN

DATE/TIME OF TEST: 01/11/2016 22:01:04

Diagnostics Check		OK	21:54
Air Blank		0.000	21:55
Control Test		0.076	21:55
Air Blank		0.000	21:55
Subject Sample	# 1	0.045	21:56
Breath Volume		2.605	21:56
Air Blank		0.000	21:57
Air Blank		0.000	21:59
Subject Sample	# 2	0.041	21:59
Breath Volume		1.882	21:59
Air Blank		0.000	22:00
Control Test		0.076	22:00
Air Blank		0.000	22:00
Diagnostics Check		OK	22:01

Data Download Date/Time: 01/31/2016 10:48

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000947

INSTRUMENT REGISTERED TO: GLADES COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BONILLA,JOSUE,D
BREATH TEST OPERATOR AGENCY: Glades County SO
ARRESTING OFFICER: ALVAREZ,VICTOR
ARRESTING OFFICER AGENCY: Glades County SO
LAST AGENCY INSPECTION: 12/31/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13314080A3
DRY GAS EXPIRY DATE: 06/05/2016
VIOLATION CODE: ADMIN
ARREST TIME: 10:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:35

SUBJECT'S LAST NAME: SANCHEZ

DATE/TIME OF TEST: 01/24/2016 11:02:15

Diagnostics Check		OK	10:57	
Air Blank		0.000	10:58	
Control Test		0.076	10:58	
Air Blank		0.000	10:58	
Subject Sample	# 1		11:01	Sequence Aborted
Breath Volume		0.101	11:01	
Air Blank		0.000	11:02	

Data Download Date/Time: 01/31/2016 10:48

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000947

INSTRUMENT REGISTERED TO: GLADES COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BONILLA,JOSUE,D
BREATH TEST OPERATOR AGENCY: Glades County SO
ARRESTING OFFICER: ALVAREZ,VICTOR
ARRESTING OFFICER AGENCY: Glades County SO
LAST AGENCY INSPECTION: 12/31/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13314080A3
DRY GAS EXPIRY DATE: 06/05/2016
VIOLATION CODE: ADMIN
ARREST TIME: 10:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:35

SUBJECT'S LAST NAME: SANCHEZ

DATE/TIME OF TEST: 01/24/2016 11:05:35

Diagnostics Check	OK	11:04	
Air Blank	0.000	11:04	
Control Test	0.072	11:04	Control Outside Tolerance
Air Blank	0.000	11:05	

Data Download Date/Time: 01/31/2016 10:48

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000947

INSTRUMENT REGISTERED TO: GLADES COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BONILLA,JOSUE,D
BREATH TEST OPERATOR AGENCY: Glades County SO
ARRESTING OFFICER: ALVAREZ,VICTOR
ARRESTING OFFICER AGENCY: Glades County SO
LAST AGENCY INSPECTION: 12/31/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13314080A3
DRY GAS EXPIRY DATE: 06/05/2016
VIOLATION CODE: ADMIN
ARREST TIME: 10:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:00

SUBJECT'S LAST NAME: SANCHEZ

DATE/TIME OF TEST: 01/24/2016 11:18:36

Diagnostics Check		OK	11:11	
Air Blank		0.000	11:11	
Control Test		0.076	11:11	
Air Blank		0.000	11:12	
Subject Sample	# 1	0.217	11:13	
Breath Volume		1.390	11:13	
Air Blank		0.000	11:14	
Air Blank		0.000	11:15	
Subject Sample	# 2	0.208	11:16	
Breath Volume		1.472	11:16	
Air Blank		0.000	11:17	
Control Test		0.074	11:17	Control Outside Tolerance
Air Blank		0.000	11:18	

Data Download Date/Time: 01/31/2016 10:48

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000948

INSTRUMENT REGISTERED TO: GLADES COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAXTER,AUDREY,D
BREATH TEST OPERATOR AGENCY: Glades County SO
ARRESTING OFFICER: ANDERSON,RICHARD
ARRESTING OFFICER AGENCY: Glades County SO
LAST AGENCY INSPECTION: 12/31/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13314080A3
DRY GAS EXPIRY DATE: 06/05/2016
VIOLATION CODE: ADMIN
ARREST TIME: 18:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:34

SUBJECT'S LAST NAME: HIGH BEAR

DATE/TIME OF TEST: 01/05/2016 19:35:51

Diagnostics Check		OK	19:26	
Air Blank		0.000	19:27	
Control Test		0.077	19:27	
Air Blank		0.000	19:28	
Subject Sample	# 1	0.280	19:28	
Breath Volume		1.156	19:28	
Air Blank		0.000	19:29	
Air Blank		0.000	19:31	
Subject Sample	# 2	0.274	19:34	Volume Not Met
Breath Volume		0.214	19:34	
Air Blank			19:35	Purge Fail
Air Blank		0.000	19:35	

Data Download Date/Time: 01/31/2016 10:44

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000948

INSTRUMENT REGISTERED TO: GLADES COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAXTER,AUDREY,D
BREATH TEST OPERATOR AGENCY: Glades County SO
ARRESTING OFFICER: ANDERSON,RICHARD
ARRESTING OFFICER AGENCY: Glades County SO
LAST AGENCY INSPECTION: 12/31/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13314080A3
DRY GAS EXPIRY DATE: 06/05/2016
VIOLATION CODE: ADMIN
ARREST TIME: 18:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:55

SUBJECT'S LAST NAME: HIGH BEAR

DATE/TIME OF TEST: 01/06/2016 05:22:20

Diagnostics Check		OK	05:15
Air Blank		0.000	05:15
Control Test		0.076	05:16
Air Blank		0.000	05:16
Subject Sample	# 1	0.177	05:16
Breath Volume		1.242	05:16
Air Blank		0.000	05:17
Air Blank		0.000	05:19
Subject Sample	# 2	0.183	05:20
Breath Volume		1.273	05:20
Air Blank		0.000	05:21
Control Test		0.077	05:21
Air Blank		0.000	05:22
Diagnostics Check		OK	05:22

Data Download Date/Time: 01/31/2016 10:44

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000948

INSTRUMENT REGISTERED TO: GLADES COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HICKS,CHRISTINA,L
BREATH TEST OPERATOR AGENCY: Glades County SO
ARRESTING OFFICER: ANDERSON,RICHARD
ARRESTING OFFICER AGENCY: Glades County SO
LAST AGENCY INSPECTION: 12/31/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13314080A3
DRY GAS EXPIRY DATE: 06/05/2016
VIOLATION CODE: ADMIN
ARREST TIME: 18:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:20

SUBJECT'S LAST NAME: HIGHBEAR

DATE/TIME OF TEST: 01/06/2016 12:06:39

Diagnostics Check		OK	12:00
Air Blank		0.000	12:00
Control Test		0.077	12:01
Air Blank		0.000	12:01
Subject Sample	# 1	0.062	12:02
Breath Volume		3.335	12:02
Air Blank		0.000	12:02
Air Blank		0.000	12:04
Subject Sample	# 2	0.059	12:05
Breath Volume		3.484	12:05
Air Blank		0.000	12:05
Control Test		0.075	12:06
Air Blank		0.000	12:06
Diagnostics Check		OK	12:06

Data Download Date/Time: 01/31/2016 10:44

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000948

INSTRUMENT REGISTERED TO: GLADES COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,JOEL,D
BREATH TEST OPERATOR AGENCY: Glades County SO
ARRESTING OFFICER: ALVEREZ,VICTOR
ARRESTING OFFICER AGENCY: Glades County SO
LAST AGENCY INSPECTION: 12/31/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13314080A3
DRY GAS EXPIRY DATE: 06/05/2016
VIOLATION CODE: ADMIN
ARREST TIME: 09:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:45

SUBJECT'S LAST NAME: SANCHEZ

DATE/TIME OF TEST: 01/24/2016 18:30:47

Diagnostics Check		OK	18:24
Air Blank		0.000	18:25
Control Test		0.077	18:25
Air Blank		0.000	18:25
Subject Sample	# 1	0.050	18:26
Breath Volume		1.324	18:26
Air Blank		0.000	18:27
Air Blank		0.000	18:28
Subject Sample	# 2	0.048	18:29
Breath Volume		1.406	18:29
Air Blank		0.000	18:29
Control Test		0.076	18:30
Air Blank		0.000	18:30
Diagnostics Check		OK	18:30

Data Download Date/Time: 01/31/2016 10:44

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000948

INSTRUMENT REGISTERED TO: GLADES COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BONILLA,JOSUE,D
BREATH TEST OPERATOR AGENCY: Glades County SO
ARRESTING OFFICER: ALVAREZ,VICTOR
ARRESTING OFFICER AGENCY: Glades County SO
LAST AGENCY INSPECTION: 12/31/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13314080A3
DRY GAS EXPIRY DATE: 06/05/2016
VIOLATION CODE: ADMIN
ARREST TIME: 08:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:40

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 01/27/2016 09:19:10

Diagnostics Check		OK	09:13
Air Blank		0.000	09:13
Control Test		0.077	09:13
Air Blank		0.000	09:14
Subject Sample	# 1	0.192	09:14
Breath Volume		2.667	09:14
Air Blank		0.000	09:15
Air Blank		0.000	09:17
Subject Sample	# 2	0.191	09:17
Breath Volume		2.988	09:17
Air Blank		0.000	09:18
Control Test		0.075	09:18
Air Blank		0.000	09:19
Diagnostics Check		OK	09:19

Data Download Date/Time: 01/31/2016 10:44

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000948

INSTRUMENT REGISTERED TO: GLADES COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BONILLA,JOSUE,D
BREATH TEST OPERATOR AGENCY: Glades County SO
ARRESTING OFFICER: ALVAREZ,VICTOR
ARRESTING OFFICER AGENCY: Glades County SO
LAST AGENCY INSPECTION: 12/31/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13314080A3
DRY GAS EXPIRY DATE: 06/05/2016
VIOLATION CODE: ADMIN
ARREST TIME: 08:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:25

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 01/27/2016 14:52:12

Diagnostics Check		OK	14:46
Air Blank		0.000	14:46
Control Test		0.077	14:46
Air Blank		0.000	14:47
Subject Sample	# 1	0.034	14:47
Breath Volume		2.707	14:47
Air Blank		0.000	14:48
Air Blank		0.000	14:50
Subject Sample	# 2	0.032	14:50
Breath Volume		2.195	14:50
Air Blank		0.000	14:51
Control Test		0.076	14:51
Air Blank		0.000	14:52
Diagnostics Check		OK	14:52

Data Download Date/Time: 01/31/2016 10:44

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000950

INSTRUMENT REGISTERED TO: HAVANA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANDREWS,DONTERIS,M
BREATH TEST OPERATOR AGENCY: Havana PD
ARRESTING OFFICER: ANDREWS,DONTERIS
ARRESTING OFFICER AGENCY: Havana PD
LAST AGENCY INSPECTION: 01/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20314080A4
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 07:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:05

SUBJECT'S LAST NAME: GIBSON

DATE/TIME OF TEST: 01/31/2016 07:35:09

Diagnostics Check		OK	07:25	
Air Blank		0.000	07:26	
Control Test		0.076	07:26	
Air Blank		0.000	07:26	
Subject Sample	# 1	0.244	07:29	
Breath Volume		1.929	07:29	
Air Blank		0.000	07:30	
Air Blank		0.000	07:31	
Subject Sample	# 2	0.232	07:33	
Breath Volume		1.683	07:33	
Air Blank		0.000	07:34	
Control Test		0.074	07:34	Control Outside Tolerance
Air Blank		0.000	07:35	

Data Download Date/Time: 02/17/2016 10:34

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000950

INSTRUMENT REGISTERED TO: HAVANA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANDREWS,DONTERIS,M
BREATH TEST OPERATOR AGENCY: Havana PD
ARRESTING OFFICER: ANDREWS,DONTERIS
ARRESTING OFFICER AGENCY: Havana PD
LAST AGENCY INSPECTION: 01/26/2016

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 20314080A4
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 07:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:05

SUBJECT'S LAST NAME: GIBSON

DATE/TIME OF TEST: 01/31/2016 07:47:03

Diagnostics Check		OK	07:39
Air Blank		0.000	07:39
Control Test		0.076	07:39
Air Blank		0.000	07:40
Subject Sample	# 1	0.234	07:42
Breath Volume		1.632	07:42
Air Blank		0.000	07:43
Air Blank		0.000	07:44
Subject Sample	# 2	0.236	07:45
Breath Volume		2.343	07:45
Air Blank		0.000	07:46
Control Test		0.076	07:46
Air Blank		0.000	07:46
Diagnostics Check		OK	07:47

Data Download Date/Time: 02/17/2016 10:34

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000953

INSTRUMENT REGISTERED TO: FRANKLIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GENTRY,JONATHAN,P
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: GENTRY,JONATHAN
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 01/27/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1841081
DRY GAS EXPIRY DATE: 04/08/2018
VIOLATION CODE: DUI
ARREST TIME: 12:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: GIPSON

DATE/TIME OF TEST: 01/29/2016 02:17:47

Diagnostics Check		OK	02:05	
Air Blank		0.000	02:06	
Control Test		0.078	02:06	
Air Blank		0.000	02:07	
Subject Sample	# 1	0.105	02:10	Volume Not Met
Breath Volume		0.015	02:10	
Air Blank		0.000	02:11	
Air Blank		0.000	02:12	
Subject Sample	# 2	0.091	02:15	Volume Not Met
Breath Volume		0.324	02:15	
Air Blank		0.000	02:16	
Control Test		0.080	02:17	
Air Blank		0.000	02:17	
Diagnostics Check		OK	02:17	

Data Download Date/Time: 02/23/2016 08:03

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000953

INSTRUMENT REGISTERED TO: FRANKLIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GENTRY,JONATHAN,P
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: GENTRY,JONATHAN
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 01/27/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1841081
DRY GAS EXPIRY DATE: 04/08/2018
VIOLATION CODE: DUI
ARREST TIME: 12:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: GIPSON

DATE/TIME OF TEST: 01/29/2016 02:32:27

Diagnostics Check		OK	02:20	
Air Blank		0.000	02:21	
Control Test		0.078	02:21	
Air Blank		0.000	02:22	
Subject Sample	# 1	0.089	02:25	Volume Not Met
Breath Volume		0.398	02:25	
Air Blank		0.000	02:26	
Air Blank		0.000	02:27	
Subject Sample	# 2	0.092	02:30	Volume Not Met
Breath Volume		0.160	02:30	
Air Blank		0.000	02:31	
Control Test		0.079	02:31	
Air Blank		0.000	02:32	
Diagnostics Check		OK	02:32	

Data Download Date/Time: 02/23/2016 08:03

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000953

INSTRUMENT REGISTERED TO: FRANKLIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GENTRY,JONATHAN,P
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: GENTRY,JONATHAN
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 01/27/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1841081
DRY GAS EXPIRY DATE: 04/08/2018
VIOLATION CODE: DUI
ARREST TIME: 12:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: GIPSON

DATE/TIME OF TEST: 01/29/2016 02:48:11

Diagnostics Check		OK	02:34	
Air Blank		0.000	02:34	
Control Test		0.078	02:35	
Air Blank		0.000	02:35	
Subject Sample	# 1	0.092	02:37	No .020 agreement
Breath Volume		2.007	02:37	
Air Blank		0.000	02:38	
Air Blank		0.000	02:40	
Subject Sample	# 2	0.079	02:43	Volume Not Met
Breath Volume		0.351	02:43	
Air Blank		0.000	02:44	
Air Blank		0.000	02:46	
Subject Sample	# 3	0.065	02:46	No .020 agreement
Breath Volume		1.875	02:46	
Air Blank		0.000	02:47	
Control Test		0.076	02:47	
Air Blank		0.000	02:48	
Diagnostics Check		OK	02:48	

Data Download Date/Time: 02/23/2016 08:03

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000955

INSTRUMENT REGISTERED TO: CLEWISTON P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,N,
BREATH TEST OPERATOR AGENCY: Clewiston PD
ARRESTING OFFICER: VANDETTI,M
ARRESTING OFFICER AGENCY: Clewiston PD
LAST AGENCY INSPECTION: 12/08/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1737087
DRY GAS EXPIRY DATE: 09/12/2017
VIOLATION CODE: DUI
ARREST TIME: 12:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:14

SUBJECT'S LAST NAME: CYPRESS

DATE/TIME OF TEST: 01/09/2016 01:38:02

Diagnostics Check		OK	01:35	
Air Blank		0.000	01:35	
Control Test		0.077	01:35	
Air Blank		0.000	01:36	
Subject Sample	# 1		01:36	Subject Test Refused
Breath Volume		0.000	01:36	
Air Blank		0.000	01:37	
Control Test		0.078	01:37	
Air Blank		0.000	01:37	
Diagnostics Check		OK	01:38	

Data Download Date/Time: 01/30/2016 20:30

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000956

INSTRUMENT REGISTERED TO: JEFFERSON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,WILLIAM,B
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: WARD,RONALD
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 12/07/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20314080A4
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 00:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: ONOKPISE

DATE/TIME OF TEST: 01/02/2016 01:17:58

Diagnostics Check		OK	01:10
Air Blank		0.000	01:11
Control Test		0.078	01:11
Air Blank		0.000	01:12
Subject Sample	# 1	0.236	01:13
Breath Volume		1.191	01:13
Air Blank		0.000	01:14
Air Blank		0.000	01:15
Subject Sample	# 2	0.240	01:16
Breath Volume		2.003	01:16
Air Blank		0.000	01:17
Control Test		0.077	01:17
Air Blank		0.000	01:17
Diagnostics Check		OK	01:17

Data Download Date/Time: 01/20/2016 09:09

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000956

INSTRUMENT REGISTERED TO: JEFFERSON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,WILLIAM,B
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: MATHEWS,DUSTIN
ARRESTING OFFICER AGENCY: Jefferson County SO
LAST AGENCY INSPECTION: 12/07/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20314080A4
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 00:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:39

SUBJECT'S LAST NAME: SIPLIN

DATE/TIME OF TEST: 01/02/2016 02:08:35

Diagnostics Check		OK	02:01
Air Blank		0.000	02:01
Control Test		0.078	02:02
Air Blank		0.000	02:02
Subject Sample	# 1	0.332	02:03
Breath Volume		1.984	02:03
Air Blank		0.000	02:03
Air Blank		0.000	02:05
Subject Sample	# 2	0.318	02:06
Breath Volume		2.613	02:06
Air Blank		0.000	02:07
Control Test		0.078	02:07
Air Blank		0.000	02:08
Diagnostics Check		OK	02:08

Data Download Date/Time: 01/20/2016 09:09

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000956

INSTRUMENT REGISTERED TO: JEFFERSON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZRAK,PAUL,M
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: MATTHEWS,DUSTIN
ARRESTING OFFICER AGENCY: Jefferson County SO
LAST AGENCY INSPECTION: 12/07/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20314080A4
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 23:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: FUTRELL

DATE/TIME OF TEST: 01/16/2016 01:03:36

Diagnostics Check		OK	01:00	
Air Blank		0.000	01:01	
Control Test		0.078	01:01	
Air Blank		0.000	01:01	
Subject Sample	# 1		01:02	Subject Test Refused
Breath Volume		0.000	01:02	
Air Blank		0.000	01:02	
Control Test		0.078	01:03	
Air Blank		0.000	01:03	
Diagnostics Check		OK	01:03	

Data Download Date/Time: 01/20/2016 09:09

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000956

INSTRUMENT REGISTERED TO: JEFFERSON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LLANES,GABRIEL,M
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: LLANES,GABRIEL
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 12/07/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20314080A4
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 04:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:10

SUBJECT'S LAST NAME: FISK

DATE/TIME OF TEST: 01/17/2016 05:37:02

Diagnostics Check		OK	05:34	
Air Blank		0.000	05:34	
Control Test		0.080	05:35	
Air Blank		0.000	05:35	
Subject Sample	# 1		05:35	Subject Test Refused
Breath Volume		0.000	05:35	
Air Blank		0.000	05:36	
Control Test		0.078	05:36	
Air Blank		0.000	05:36	
Diagnostics Check		OK	05:37	

Data Download Date/Time: 01/20/2016 09:09

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000956

INSTRUMENT REGISTERED TO: JEFFERSON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BONTRAGER,WILLIAM,L
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: HANNERS,SHERI
ARRESTING OFFICER AGENCY: Jefferson County SO
LAST AGENCY INSPECTION: 12/07/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20314080A4
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 21:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: ROBINSON

DATE/TIME OF TEST: 01/18/2016 23:13:01

Diagnostics Check		OK	23:01	
Air Blank		0.000	23:01	
Control Test		0.079	23:02	
Air Blank		0.000	23:02	
Subject Sample	# 1	0.114	23:05	Volume Not Met
Breath Volume		0.070	23:05	
Air Blank		0.000	23:06	
Air Blank		0.000	23:08	
Subject Sample	# 2	0.110	23:11	Volume Not Met
Breath Volume		0.066	23:11	
Air Blank		0.000	23:12	
Control Test		0.079	23:12	
Air Blank		0.000	23:12	
Diagnostics Check		OK	23:13	

Data Download Date/Time: 01/20/2016 09:09

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000956

INSTRUMENT REGISTERED TO: JEFFERSON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOATWRIGHT,MATTHEW,C
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: DEAS,BRANDYN
ARRESTING OFFICER AGENCY: Jefferson County SO
LAST AGENCY INSPECTION: 01/19/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20314080A4
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 03:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:39

SUBJECT'S LAST NAME: PAYNE

DATE/TIME OF TEST: 01/30/2016 04:10:32

Diagnostics Check		OK	04:04
Air Blank		0.000	04:04
Control Test		0.078	04:05
Air Blank		0.000	04:05
Subject Sample	# 1	0.129	04:06
Breath Volume		1.750	04:06
Air Blank		0.000	04:06
Air Blank		0.000	04:08
Subject Sample	# 2	0.117	04:09
Breath Volume		1.410	04:09
Air Blank		0.000	04:09
Control Test		0.078	04:09
Air Blank		0.000	04:10
Diagnostics Check		OK	04:10

Data Download Date/Time: 02/04/2016 11:14

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHULZE,JON,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: SCHULZE,JON
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 12/10/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:49

SUBJECT'S LAST NAME: SKINNER

DATE/TIME OF TEST: 01/01/2016 03:17:29

Diagnostics Check		OK	03:10
Air Blank		0.000	03:11
Control Test		0.080	03:11
Air Blank		0.000	03:12
Subject Sample	# 1	0.150	03:12
Breath Volume		2.250	03:12
Air Blank		0.000	03:13
Air Blank		0.000	03:15
Subject Sample	# 2	0.165	03:15
Breath Volume		3.183	03:15
Air Blank		0.000	03:16
Control Test		0.079	03:16
Air Blank		0.000	03:17
Diagnostics Check		OK	03:17

Data Download Date/Time: 01/25/2016 10:47

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROY,G,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: ROY,G
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 12/10/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:43

SUBJECT'S LAST NAME: HAGEN

DATE/TIME OF TEST: 01/02/2016 02:23:44

Diagnostics Check		OK	02:16
Air Blank		0.000	02:17
Control Test		0.080	02:17
Air Blank		0.000	02:17
Subject Sample	# 1	0.198	02:19
Breath Volume		2.457	02:19
Air Blank		0.000	02:19
Air Blank		0.000	02:21
Subject Sample	# 2	0.193	02:22
Breath Volume		2.402	02:22
Air Blank		0.000	02:22
Control Test		0.080	02:23
Air Blank		0.000	02:23
Diagnostics Check		OK	02:23

Data Download Date/Time: 01/25/2016 10:47

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CUTCLIFFE,MARY,P
BREATH TEST OPERATOR AGENCY: Leon County SO
ARRESTING OFFICER: CUTCLIFFE,MARY
ARRESTING OFFICER AGENCY: Leon County SO
LAST AGENCY INSPECTION: 12/10/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:17

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 01/07/2016 23:42:08

Diagnostics Check		OK	23:39	
Air Blank		0.000	23:39	
Control Test		0.081	23:40	
Air Blank		0.000	23:40	
Subject Sample	# 1		23:40	Subject Test Refused
Breath Volume		0.000	23:40	
Air Blank		0.000	23:41	
Control Test		0.081	23:41	
Air Blank		0.000	23:42	
Diagnostics Check		OK	23:42	

Data Download Date/Time: 01/25/2016 10:47

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROY,G,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: ROY,G
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 12/10/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:22

SUBJECT'S LAST NAME: BETTNER

DATE/TIME OF TEST: 01/09/2016 03:51:30

Diagnostics Check		OK	03:45
Air Blank		0.000	03:45
Control Test		0.080	03:46
Air Blank		0.000	03:46
Subject Sample	# 1	0.184	03:47
Breath Volume		1.484	03:47
Air Blank		0.000	03:47
Air Blank		0.000	03:49
Subject Sample	# 2	0.180	03:50
Breath Volume		1.449	03:50
Air Blank		0.000	03:50
Control Test		0.080	03:50
Air Blank		0.000	03:51
Diagnostics Check		OK	03:51

Data Download Date/Time: 01/25/2016 10:47

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMIDT,M,D
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: SMIDT,M
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 12/10/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:15

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 01/09/2016 04:49:42

Diagnostics Check		OK	04:43
Air Blank		0.000	04:43
Control Test		0.080	04:44
Air Blank		0.000	04:44
Subject Sample	# 1	0.189	04:45
Breath Volume		2.289	04:45
Air Blank		0.000	04:45
Air Blank		0.000	04:47
Subject Sample	# 2	0.180	04:48
Breath Volume		2.996	04:48
Air Blank		0.000	04:48
Control Test		0.079	04:49
Air Blank		0.000	04:49
Diagnostics Check		OK	04:49

Data Download Date/Time: 01/25/2016 10:47

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILSON,BRANDON,J
BREATH TEST OPERATOR AGENCY: FL State University
ARRESTING OFFICER: WILSON,BRANDON
ARRESTING OFFICER AGENCY: FL State University
LAST AGENCY INSPECTION: 12/10/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:54

SUBJECT'S LAST NAME: EZELL

DATE/TIME OF TEST: 01/10/2016 02:36:02

Diagnostics Check		OK	02:29
Air Blank		0.000	02:30
Control Test		0.080	02:30
Air Blank		0.000	02:31
Subject Sample	# 1	0.218	02:31
Breath Volume		3.250	02:31
Air Blank		0.000	02:32
Air Blank		0.000	02:34
Subject Sample	# 2	0.207	02:34
Breath Volume		1.441	02:34
Air Blank		0.000	02:35
Control Test		0.078	02:35
Air Blank		0.000	02:35
Diagnostics Check		OK	02:36

Data Download Date/Time: 01/25/2016 10:47

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRAXTON,D,L
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: BRAXTON,D
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 12/10/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 01/12/2016 04:15:35

Diagnostics Check		OK	04:08
Air Blank		0.000	04:09
Control Test		0.081	04:09
Air Blank		0.000	04:09
Subject Sample	# 1	0.144	04:10
Breath Volume		1.871	04:10
Air Blank		0.000	04:11
Air Blank		0.000	04:13
Subject Sample	# 2	0.131	04:13
Breath Volume		1.683	04:13
Air Blank		0.000	04:14
Control Test		0.078	04:14
Air Blank		0.000	04:15
Diagnostics Check		OK	04:15

Data Download Date/Time: 01/25/2016 10:47

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FINUFF,DANIEL,W
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: FINUFF,DANIEL
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 12/10/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:32

SUBJECT'S LAST NAME: KENNEDY

DATE/TIME OF TEST: 01/13/2016 23:00:20

Diagnostics Check		OK	22:52
Air Blank		0.000	22:53
Control Test		0.081	22:53
Air Blank		0.000	22:53
Subject Sample	# 1	0.147	22:54
Breath Volume		2.011	22:54
Air Blank		0.000	22:55
Air Blank		0.000	22:57
Subject Sample	# 2	0.147	22:58
Breath Volume		2.894	22:58
Air Blank		0.000	22:59
Control Test		0.080	22:59
Air Blank		0.000	23:00
Diagnostics Check		OK	23:00

Data Download Date/Time: 01/25/2016 10:47

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHULZE,JON,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: SCHULZE,JON
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 12/10/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:56

SUBJECT'S LAST NAME: NELSON

DATE/TIME OF TEST: 01/15/2016 01:23:05

Diagnostics Check		OK	01:16
Air Blank		0.000	01:17
Control Test		0.080	01:17
Air Blank		0.000	01:18
Subject Sample	# 1	0.155	01:18
Breath Volume		1.550	01:18
Air Blank		0.000	01:19
Air Blank		0.000	01:21
Subject Sample	# 2	0.156	01:21
Breath Volume		1.531	01:21
Air Blank		0.000	01:22
Control Test		0.080	01:22
Air Blank		0.000	01:22
Diagnostics Check		OK	01:23

Data Download Date/Time: 01/25/2016 10:47

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORAN,CORY,P
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: MORAN,CORY
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 12/10/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: CARRAWAY

DATE/TIME OF TEST: 01/18/2016 04:14:46

Diagnostics Check		OK	04:07
Air Blank		0.000	04:08
Control Test		0.080	04:08
Air Blank		0.000	04:08
Subject Sample	# 1	0.187	04:09
Breath Volume		1.230	04:09
Air Blank		0.000	04:10
Air Blank		0.000	04:12
Subject Sample	# 2	0.179	04:13
Breath Volume		1.335	04:13
Air Blank		0.000	04:13
Control Test		0.080	04:14
Air Blank		0.000	04:14
Diagnostics Check		OK	04:14

Data Download Date/Time: 01/25/2016 10:47

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORAN,CORY,P
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: MORAN,CORY
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 12/10/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: BUNCH

DATE/TIME OF TEST: 01/22/2016 02:32:05

Diagnostics Check		OK	02:29	
Air Blank		0.000	02:29	
Control Test		0.080	02:30	
Air Blank		0.000	02:30	
Subject Sample	# 1		02:30	Subject Test Refused
Breath Volume		0.000	02:30	
Air Blank		0.000	02:31	
Control Test		0.080	02:31	
Air Blank		0.000	02:32	
Diagnostics Check		OK	02:32	

Data Download Date/Time: 01/25/2016 10:47

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRELL,CASEY,B
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: HARRELL,CASEY
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 12/10/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: LONG

DATE/TIME OF TEST: 01/23/2016 23:38:45

Diagnostics Check		OK	23:32
Air Blank		0.000	23:33
Control Test		0.080	23:33
Air Blank		0.000	23:33
Subject Sample	# 1	0.205	23:34
Breath Volume		3.003	23:34
Air Blank		0.000	23:35
Air Blank		0.000	23:36
Subject Sample	# 2	0.203	23:37
Breath Volume		2.984	23:37
Air Blank		0.000	23:37
Control Test		0.079	23:38
Air Blank		0.000	23:38
Diagnostics Check		OK	23:38

Data Download Date/Time: 01/25/2016 10:47

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRIS,JAMES,J
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: HARRIS,JAMES
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 12/10/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: ETIENNE

DATE/TIME OF TEST: 01/24/2016 04:29:30

Diagnostics Check		OK	04:21
Air Blank		0.000	04:22
Control Test		0.081	04:22
Air Blank		0.000	04:22
Subject Sample	# 1	0.232	04:25
Breath Volume		1.265	04:25
Air Blank		0.000	04:25
Air Blank		0.000	04:27
Subject Sample	# 2	0.237	04:27
Breath Volume		1.386	04:27
Air Blank		0.000	04:28
Control Test		0.079	04:28
Air Blank		0.000	04:29
Diagnostics Check		OK	04:29

Data Download Date/Time: 01/25/2016 10:47

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILSON,BRANDON,J
BREATH TEST OPERATOR AGENCY: FL State University
ARRESTING OFFICER: WILSON,BRANDON
ARRESTING OFFICER AGENCY: FL State University
LAST AGENCY INSPECTION: 01/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: BAKER

DATE/TIME OF TEST: 01/26/2016 03:26:41

Diagnostics Check		OK	03:20
Air Blank		0.000	03:20
Control Test		0.080	03:21
Air Blank		0.000	03:21
Subject Sample	# 1	0.148	03:22
Breath Volume		1.691	03:22
Air Blank		0.000	03:22
Air Blank		0.000	03:24
Subject Sample	# 2	0.152	03:25
Breath Volume		2.667	03:25
Air Blank		0.000	03:25
Control Test		0.081	03:26
Air Blank		0.000	03:26
Diagnostics Check		OK	03:26

Data Download Date/Time: 02/09/2016 10:54

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KELLER,DAVID,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: KELLER,D
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 01/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 05:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:42

SUBJECT'S LAST NAME: BROGDEN

DATE/TIME OF TEST: 01/26/2016 06:08:30

Diagnostics Check		OK	06:02
Air Blank		0.000	06:02
Control Test		0.080	06:02
Air Blank		0.000	06:03
Subject Sample	# 1	0.216	06:04
Breath Volume		3.507	06:04
Air Blank		0.000	06:04
Air Blank		0.000	06:06
Subject Sample	# 2	0.210	06:06
Breath Volume		3.117	06:06
Air Blank		0.000	06:07
Control Test		0.080	06:07
Air Blank		0.000	06:08
Diagnostics Check		OK	06:08

Data Download Date/Time: 02/09/2016 10:54

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KELLER,D,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: KELLER,D
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 01/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:43

SUBJECT'S LAST NAME: JOHNSTON

DATE/TIME OF TEST: 01/26/2016 23:13:46

Diagnostics Check		OK	23:03
Air Blank		0.000	23:04
Control Test		0.080	23:04
Air Blank		0.000	23:05
Subject Sample	# 1	0.223	23:07
Breath Volume		1.460	23:07
Air Blank		0.000	23:08
Air Blank		0.000	23:10
Subject Sample	# 2	0.226	23:12
Breath Volume		2.421	23:12
Air Blank		0.000	23:12
Control Test		0.079	23:13
Air Blank		0.000	23:13
Diagnostics Check		OK	23:13

Data Download Date/Time: 02/09/2016 10:54

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILSON,BRANDON,J
BREATH TEST OPERATOR AGENCY: FL State University
ARRESTING OFFICER: WILSON,BRANDON
ARRESTING OFFICER AGENCY: FL State University
LAST AGENCY INSPECTION: 01/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:47

SUBJECT'S LAST NAME: BRINSON

DATE/TIME OF TEST: 01/30/2016 04:15:35

Diagnostics Check		OK	04:12	
Air Blank		0.000	04:13	
Control Test		0.080	04:13	
Air Blank		0.000	04:14	
Subject Sample	# 1		04:14	Subject Test Refused
Breath Volume		0.000	04:14	
Air Blank		0.000	04:14	
Control Test		0.081	04:15	
Air Blank		0.000	04:15	
Diagnostics Check		OK	04:15	

Data Download Date/Time: 02/09/2016 10:54

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,JEFFREY,A
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: THOMPSON,JEFFREY
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 01/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: TRENT

DATE/TIME OF TEST: 01/31/2016 04:27:16

Diagnostics Check		OK	04:21
Air Blank		0.000	04:21
Control Test		0.079	04:21
Air Blank		0.000	04:22
Subject Sample	# 1	0.273	04:22
Breath Volume		1.976	04:22
Air Blank		0.000	04:23
Air Blank		0.000	04:25
Subject Sample	# 2	0.277	04:25
Breath Volume		1.226	04:25
Air Blank		0.000	04:26
Control Test		0.078	04:26
Air Blank		0.000	04:27
Diagnostics Check		OK	04:27

Data Download Date/Time: 02/09/2016 10:54

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHAW,EMILY,P
BREATH TEST OPERATOR AGENCY: Leon County SO
ARRESTING OFFICER: BATEMAN,JOSEPH
ARRESTING OFFICER AGENCY: Leon County SO
LAST AGENCY INSPECTION: 12/10/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:15

SUBJECT'S LAST NAME: OGDEN

DATE/TIME OF TEST: 01/01/2016 06:04:13

Diagnostics Check		OK	05:57
Air Blank		0.000	05:58
Control Test		0.080	05:58
Air Blank		0.000	05:59
Subject Sample	# 1	0.159	05:59
Breath Volume		1.699	05:59
Air Blank		0.000	06:00
Air Blank		0.000	06:02
Subject Sample	# 2	0.170	06:02
Breath Volume		2.351	06:02
Air Blank		0.000	06:03
Control Test		0.079	06:03
Air Blank		0.000	06:04
Diagnostics Check		OK	06:04

Data Download Date/Time: 01/25/2016 10:44

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUSCITTO,C,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: RUSCITTO,C
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 12/10/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 19:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:07

SUBJECT'S LAST NAME: CONDRY

DATE/TIME OF TEST: 01/04/2016 20:38:26

Diagnostics Check		OK	20:32
Air Blank		0.000	20:32
Control Test		0.080	20:32
Air Blank		0.000	20:33
Subject Sample	# 1	0.101	20:33
Breath Volume		2.792	20:33
Air Blank		0.000	20:34
Air Blank		0.000	20:36
Subject Sample	# 2	0.100	20:36
Breath Volume		2.601	20:36
Air Blank		0.000	20:37
Control Test		0.079	20:37
Air Blank		0.000	20:38
Diagnostics Check		OK	20:38

Data Download Date/Time: 01/25/2016 10:44

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHULZE,JON,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: SCHULZE,JON
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 12/10/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:51

SUBJECT'S LAST NAME: COGBURN

DATE/TIME OF TEST: 01/08/2016 01:18:47

Diagnostics Check		OK	01:11
Air Blank		0.000	01:12
Control Test		0.079	01:12
Air Blank		0.000	01:13
Subject Sample	# 1	0.076	01:14
Breath Volume		3.101	01:14
Air Blank		0.000	01:14
Air Blank		0.000	01:16
Subject Sample	# 2	0.073	01:17
Breath Volume		2.699	01:17
Air Blank		0.000	01:17
Control Test		0.079	01:18
Air Blank		0.000	01:18
Diagnostics Check		OK	01:18

Data Download Date/Time: 01/25/2016 10:44

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILSON,BRANDON,J
BREATH TEST OPERATOR AGENCY: FL State University
ARRESTING OFFICER: WILSON,BRANDON
ARRESTING OFFICER AGENCY: FL State University
LAST AGENCY INSPECTION: 12/10/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 01/15/2016 03:03:58

Diagnostics Check		OK	02:58	
Air Blank		0.000	02:59	
Control Test		0.080	02:59	
Air Blank		0.000	03:00	
Subject Sample	# 1		03:02	Subject Test Refused
Breath Volume		0.000	03:02	
Air Blank		0.000	03:03	
Control Test		0.077	03:03	
Air Blank		0.000	03:03	
Diagnostics Check		OK	03:03	

Data Download Date/Time: 01/25/2016 10:44

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHULZE,JON,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: SCHULZE,JON
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 12/10/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: FAIRCLOTH

DATE/TIME OF TEST: 01/17/2016 01:27:22

Diagnostics Check		OK	01:20
Air Blank		0.000	01:21
Control Test		0.079	01:21
Air Blank		0.000	01:21
Subject Sample	# 1	0.261	01:22
Breath Volume		1.304	01:22
Air Blank		0.000	01:23
Air Blank		0.000	01:25
Subject Sample	# 2	0.275	01:25
Breath Volume		1.781	01:25
Air Blank		0.000	01:26
Control Test		0.080	01:26
Air Blank		0.000	01:27
Diagnostics Check		OK	01:27

Data Download Date/Time: 01/25/2016 10:44

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KENNEDY,CATHY,A
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: BURCH,WESLEY
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 12/10/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 10:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:52

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 01/17/2016 12:23:44

Diagnostics Check		OK	12:17
Air Blank		0.000	12:18
Control Test		0.079	12:18
Air Blank		0.000	12:18
Subject Sample	# 1	0.145	12:19
Breath Volume		1.589	12:19
Air Blank		0.000	12:20
Air Blank		0.000	12:21
Subject Sample	# 2	0.135	12:22
Breath Volume		2.265	12:22
Air Blank		0.000	12:22
Control Test		0.078	12:23
Air Blank		0.000	12:23
Diagnostics Check		OK	12:23

Data Download Date/Time: 01/25/2016 10:44

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROY,G,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: ROY,G
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 12/10/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: CHIKUJI

DATE/TIME OF TEST: 01/22/2016 02:21:41

Diagnostics Check		OK	02:13	
Air Blank		0.000	02:13	
Control Test		0.079	02:14	
Air Blank		0.000	02:14	
Subject Sample	# 1	0.164	02:17	Volume Not Met
Breath Volume		0.097	02:17	
Air Blank		0.000	02:18	
Air Blank		0.000	02:20	
Subject Sample	# 2		02:20	Subject Test Refused
Breath Volume		0.000	02:20	
Air Blank		0.000	02:20	
Control Test		0.077	02:21	
Air Blank		0.000	02:21	
Diagnostics Check		OK	02:21	

Data Download Date/Time: 01/25/2016 10:44

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COMBS,M,R
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: AULT,S
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 12/10/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 16:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:42

SUBJECT'S LAST NAME: ARREGUIN

DATE/TIME OF TEST: 01/22/2016 17:16:07

Diagnostics Check		OK	17:09
Air Blank		0.000	17:09
Control Test		0.080	17:10
Air Blank		0.000	17:10
Subject Sample	# 1	0.199	17:11
Breath Volume		3.621	17:11
Air Blank		0.000	17:12
Air Blank		0.000	17:13
Subject Sample	# 2	0.199	17:14
Breath Volume		2.792	17:14
Air Blank		0.000	17:15
Control Test		0.078	17:15
Air Blank		0.000	17:16
Diagnostics Check		OK	17:16

Data Download Date/Time: 01/25/2016 10:44

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRIS,JAMES,J
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: HARRIS,JAMES
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 12/10/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:50

SUBJECT'S LAST NAME: BUSBY

DATE/TIME OF TEST: 01/23/2016 05:17:22

Diagnostics Check		OK	05:10
Air Blank		0.000	05:11
Control Test		0.080	05:11
Air Blank		0.000	05:11
Subject Sample	# 1	0.117	05:12
Breath Volume		1.292	05:12
Air Blank		0.000	05:13
Air Blank		0.000	05:15
Subject Sample	# 2	0.126	05:15
Breath Volume		1.605	05:15
Air Blank		0.000	05:16
Control Test		0.080	05:16
Air Blank		0.000	05:17
Diagnostics Check		OK	05:17

Data Download Date/Time: 01/25/2016 10:44

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BONTRAGER,WILLIAM,L
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: BONTRAGER,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 12/10/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 05:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:30

SUBJECT'S LAST NAME: MARDER

DATE/TIME OF TEST: 01/23/2016 06:24:34

Diagnostics Check		OK	06:17
Air Blank		0.000	06:17
Control Test		0.080	06:18
Air Blank		0.000	06:18
Subject Sample	# 1	0.122	06:19
Breath Volume		1.890	06:19
Air Blank		0.000	06:20
Air Blank		0.000	06:22
Subject Sample	# 2	0.117	06:23
Breath Volume		1.335	06:23
Air Blank		0.000	06:23
Control Test		0.079	06:23
Air Blank		0.000	06:24
Diagnostics Check		OK	06:24

Data Download Date/Time: 01/25/2016 10:44

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FINUFF,DANIEL,W
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: FINUFF,DANIEL
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 12/10/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: ROLLINS

DATE/TIME OF TEST: 01/24/2016 03:03:17

Diagnostics Check		OK	02:56
Air Blank		0.000	02:56
Control Test		0.080	02:56
Air Blank		0.000	02:57
Subject Sample	# 1	0.065	02:58
Breath Volume		1.132	02:58
Air Blank		0.000	02:59
Air Blank		0.000	03:01
Subject Sample	# 2	0.067	03:01
Breath Volume		1.855	03:01
Air Blank		0.000	03:02
Control Test		0.080	03:02
Air Blank		0.000	03:03
Diagnostics Check		OK	03:03

Data Download Date/Time: 01/25/2016 10:44

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,PARISE,L
BREATH TEST OPERATOR AGENCY: FL State University
ARRESTING OFFICER: NATIONS,OMA
ARRESTING OFFICER AGENCY: FL State University
LAST AGENCY INSPECTION: 12/10/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:01

SUBJECT'S LAST NAME: PATEL

DATE/TIME OF TEST: 01/24/2016 03:35:29

Diagnostics Check		OK	03:25	
Air Blank		0.000	03:25	
Control Test		0.079	03:26	
Air Blank		0.000	03:26	
Subject Sample	# 1		03:29	No Sample Provided
Breath Volume		0.000	03:29	
Air Blank		0.000	03:30	
Air Blank		0.000	03:32	
Subject Sample	# 2		03:35	Sequence Aborted
Breath Volume		0.000	03:35	
Air Blank		0.000	03:35	

Data Download Date/Time: 01/25/2016 10:44

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,PARISE,L
BREATH TEST OPERATOR AGENCY: FL State University
ARRESTING OFFICER: NATION,OMA
ARRESTING OFFICER AGENCY: FL State University
LAST AGENCY INSPECTION: 12/10/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:01

SUBJECT'S LAST NAME: PATEL

DATE/TIME OF TEST: 01/24/2016 03:43:53

Diagnostics Check		OK	03:40	
Air Blank		0.000	03:41	
Control Test		0.080	03:41	
Air Blank		0.000	03:41	
Subject Sample	# 1		03:42	Subject Test Refused
Breath Volume		0.000	03:42	
Air Blank		0.000	03:42	
Control Test		0.078	03:43	
Air Blank		0.000	03:43	
Diagnostics Check		OK	03:43	

Data Download Date/Time: 01/25/2016 10:44

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORAN,CORY,P
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: MORAN,CORY
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 12/10/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: KIRK

DATE/TIME OF TEST: 01/24/2016 04:29:35

Diagnostics Check		OK	04:23
Air Blank		0.000	04:23
Control Test		0.080	04:23
Air Blank		0.000	04:24
Subject Sample	# 1	0.117	04:25
Breath Volume		1.511	04:25
Air Blank		0.000	04:25
Air Blank		0.000	04:27
Subject Sample	# 2	0.112	04:28
Breath Volume		2.195	04:28
Air Blank		0.000	04:28
Control Test		0.080	04:28
Air Blank		0.000	04:29
Diagnostics Check		OK	04:29

Data Download Date/Time: 01/25/2016 10:44

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CUTCLIFFE,BRIAN,R
BREATH TEST OPERATOR AGENCY: FL State University
ARRESTING OFFICER: CUTCLIFFE,BRIAN
ARRESTING OFFICER AGENCY: FL State University
LAST AGENCY INSPECTION: 01/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:54

SUBJECT'S LAST NAME: KRUTZA

DATE/TIME OF TEST: 01/27/2016 02:26:06

Diagnostics Check		OK	02:16
Air Blank		0.000	02:17
Control Test		0.079	02:17
Air Blank		0.000	02:17
Subject Sample	# 1	0.193	02:18
Breath Volume		3.000	02:18
Air Blank		0.000	02:19
Air Blank		0.000	02:21
Subject Sample	# 2	0.170	02:21
Breath Volume		2.148	02:21
Air Blank		0.000	02:22
Air Blank		0.000	02:24
Subject Sample	# 3	0.183	02:24
Breath Volume		2.359	02:24
Air Blank		0.000	02:25
Control Test		0.076	02:25
Air Blank		0.000	02:26
Diagnostics Check		OK	02:26

Data Download Date/Time: 02/09/2016 10:49

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAINT JUSTE,MILTON,
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: SAINT JUSTE,MILTON
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 01/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: ICAL YALIBAT

DATE/TIME OF TEST: 01/31/2016 03:50:14

Diagnostics Check		OK	03:38	
Air Blank		0.000	03:39	
Control Test		0.080	03:39	
Air Blank		0.000	03:39	
Subject Sample	# 1	0.184	03:42	Volume Not Met
Breath Volume		0.984	03:42	
Air Blank		0.000	03:43	
Air Blank		0.000	03:45	
Subject Sample	# 2	0.219	03:48	Slope Not Level
Breath Volume		1.113	03:48	
Air Blank		0.000	03:49	
Control Test		0.077	03:49	
Air Blank		0.000	03:50	
Diagnostics Check		OK	03:50	

Data Download Date/Time: 02/09/2016 10:49

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CUTCLIFFE,BRIAN,R
BREATH TEST OPERATOR AGENCY: FL State University
ARRESTING OFFICER: CUTCLIFFE,BRIAN
ARRESTING OFFICER AGENCY: FL State University
LAST AGENCY INSPECTION: 01/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:18

SUBJECT'S LAST NAME: RIVERS

DATE/TIME OF TEST: 01/31/2016 04:48:10

Diagnostics Check		OK	04:38
Air Blank		0.000	04:39
Control Test		0.079	04:39
Air Blank		0.000	04:40
Subject Sample	# 1	0.197	04:42
Breath Volume		1.289	04:42
Air Blank		0.000	04:43
Air Blank		0.000	04:44
Subject Sample	# 2	0.192	04:46
Breath Volume		1.324	04:46
Air Blank		0.000	04:47
Control Test		0.076	04:47
Air Blank		0.000	04:48
Diagnostics Check		OK	04:48

Data Download Date/Time: 02/09/2016 10:49

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000960

INSTRUMENT REGISTERED TO: LIBERTY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GENTRY,JONATHAN,P
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: GENTRY,JONATHAN
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 01/08/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20314080A4
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 21:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:28

SUBJECT'S LAST NAME: GRAHAM

DATE/TIME OF TEST: 01/12/2016 22:56:43

Diagnostics Check		OK	22:50
Air Blank		0.000	22:50
Control Test		0.081	22:50
Air Blank		0.000	22:51
Subject Sample	# 1	0.123	22:51
Breath Volume		3.089	22:51
Air Blank		0.000	22:52
Air Blank		0.000	22:54
Subject Sample	# 2	0.118	22:55
Breath Volume		3.148	22:55
Air Blank		0.000	22:55
Control Test		0.082	22:56
Air Blank		0.000	22:56
Diagnostics Check		OK	22:56

Data Download Date/Time: 02/18/2016 09:50

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000960

INSTRUMENT REGISTERED TO: LIBERTY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHEPARD,JOSHUA,M
BREATH TEST OPERATOR AGENCY: Liberty County SO
ARRESTING OFFICER: BIGGS,JOSHUA
ARRESTING OFFICER AGENCY: Liberty County SO
LAST AGENCY INSPECTION: 01/08/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20314080A4
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 22:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:52

SUBJECT'S LAST NAME: CURLING

DATE/TIME OF TEST: 01/31/2016 23:44:27

Diagnostics Check		OK	23:38
Air Blank		0.000	23:38
Control Test		0.082	23:38
Air Blank		0.000	23:39
Subject Sample	# 1	0.130	23:39
Breath Volume		2.453	23:39
Air Blank		0.000	23:40
Air Blank		0.000	23:42
Subject Sample	# 2	0.131	23:42
Breath Volume		2.402	23:42
Air Blank		0.000	23:43
Control Test		0.082	23:43
Air Blank		0.000	23:44
Diagnostics Check		OK	23:44

Data Download Date/Time: 02/18/2016 09:50

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000962

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CROSS, SR,MICHAEL,S
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: CROSS, SR,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 12/02/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20314080A4
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 00:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: CABRERA

DATE/TIME OF TEST: 01/01/2016 00:37:37

Diagnostics Check		OK	00:35	
Air Blank		0.000	00:36	
Control Test		0.079	00:36	
Air Blank		0.000	00:36	
Subject Sample	# 1		00:37	Sequence Aborted
Breath Volume		0.000	00:37	
Air Blank		0.000	00:37	

Data Download Date/Time: 01/21/2016 10:40

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000965

INSTRUMENT REGISTERED TO: QUINCY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GENTRY,JONATHAN,P
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: GENTRY,JONATHAN
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 12/01/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20314080A4
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 17:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:37

SUBJECT'S LAST NAME: WYATT

DATE/TIME OF TEST: 01/11/2016 19:03:55

Diagnostics Check		OK	18:57	
Air Blank		0.000	18:58	
Control Test		0.080	18:58	
Air Blank		0.000	18:59	
Subject Sample	# 1	0.101	19:02	Volume Not Met
Breath Volume		0.164	19:02	
Air Blank			19:03	Purge Fail
Air Blank		0.000	19:03	

Data Download Date/Time: 01/15/2016 06:45

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000965

INSTRUMENT REGISTERED TO: QUINCY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GENTRY,JONATHAN,P
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: GENTRY,JONATHAN
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 12/01/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20314080A4
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 17:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:37

SUBJECT'S LAST NAME: WYATT

DATE/TIME OF TEST: 01/11/2016 19:15:04

Diagnostics Check		OK	19:06	
Air Blank		0.000	19:06	
Control Test		0.080	19:07	
Air Blank		0.000	19:07	
Subject Sample	# 1	0.151	19:08	
Breath Volume		1.636	19:08	
Air Blank		0.000	19:08	
Air Blank		0.000	19:10	
Subject Sample	# 2	0.148	19:13	Volume Not Met
Breath Volume		0.042	19:13	
Air Blank			19:14	Purge Fail
Air Blank		0.000	19:15	

Data Download Date/Time: 01/15/2016 06:45

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000965

INSTRUMENT REGISTERED TO: QUINCY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GENTRY,JONATHAN,P
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: GENTRY,JONATHAN
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 12/01/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20314080A4
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 17:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:37

SUBJECT'S LAST NAME: WYATT

DATE/TIME OF TEST: 01/11/2016 19:33:56

Diagnostics Check		OK	19:26
Air Blank		0.000	19:27
Control Test		0.079	19:27
Air Blank		0.000	19:28
Subject Sample	# 1	0.138	19:29
Breath Volume		1.195	19:29
Air Blank		0.000	19:30
Air Blank		0.000	19:32
Subject Sample	# 2	0.120	19:32
Breath Volume		1.136	19:32
Air Blank		0.000	19:33
Control Test		0.078	19:33
Air Blank		0.000	19:33
Diagnostics Check		OK	19:33

Data Download Date/Time: 01/15/2016 06:45

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANTHONY,MARK,A
BREATH TEST OPERATOR AGENCY: Kissimmee PD
ARRESTING OFFICER: ANTHONY,MARK
ARRESTING OFFICER AGENCY: Kissimmee PD
LAST AGENCY INSPECTION: 12/03/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1578837
DRY GAS EXPIRY DATE: 10/16/2016
VIOLATION CODE: DUI
ARREST TIME: 23:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: VAN KLEEF

DATE/TIME OF TEST: 01/01/2016 00:46:22

Diagnostics Check		OK	00:40
Air Blank		0.000	00:40
Control Test		0.080	00:41
Air Blank		0.000	00:41
Subject Sample	# 1	0.012	00:42
Breath Volume		1.222	00:42
Air Blank		0.000	00:42
Air Blank		0.000	00:44
Subject Sample	# 2	0.010	00:45
Breath Volume		1.902	00:45
Air Blank		0.000	00:45
Control Test		0.079	00:45
Air Blank		0.000	00:46
Diagnostics Check		OK	00:46

Data Download Date/Time: 01/05/2016 10:48

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LYONS,D,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: WAGNER,W
ARRESTING OFFICER AGENCY: St. Cloud PD
LAST AGENCY INSPECTION: 12/03/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1578837
DRY GAS EXPIRY DATE: 10/16/2016
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: SANKAR

DATE/TIME OF TEST: 01/02/2016 01:22:04

Diagnostics Check		OK	01:16
Air Blank		0.000	01:16
Control Test		0.079	01:16
Air Blank		0.000	01:17
Subject Sample	# 1	0.158	01:17
Breath Volume		1.925	01:17
Air Blank		0.000	01:18
Air Blank		0.000	01:20
Subject Sample	# 2	0.152	01:20
Breath Volume		1.859	01:20
Air Blank		0.000	01:21
Control Test		0.079	01:21
Air Blank		0.000	01:21
Diagnostics Check		OK	01:22

Data Download Date/Time: 01/05/2016 10:48

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KLOZNICK,ROBERT,J
BREATH TEST OPERATOR AGENCY: Kissimmee PD
ARRESTING OFFICER: KLOZNICK,ROBERT
ARRESTING OFFICER AGENCY: Kissimmee PD
LAST AGENCY INSPECTION: 12/03/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1578837
DRY GAS EXPIRY DATE: 10/16/2016
VIOLATION CODE: DUI
ARREST TIME: 03:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:29

SUBJECT'S LAST NAME: MANCERA CARDENAS

DATE/TIME OF TEST: 01/02/2016 04:31:10

Diagnostics Check		OK	04:24
Air Blank		0.000	04:25
Control Test		0.079	04:25
Air Blank		0.000	04:25
Subject Sample	# 1	0.244	04:26
Breath Volume		1.835	04:26
Air Blank		0.000	04:27
Air Blank		0.000	04:29
Subject Sample	# 2	0.252	04:29
Breath Volume		1.714	04:29
Air Blank		0.000	04:30
Control Test		0.079	04:30
Air Blank		0.000	04:31
Diagnostics Check		OK	04:31

Data Download Date/Time: 01/05/2016 10:48

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRUZ,AYLER,
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: CRUZ,AYLER
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 01/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1578837
DRY GAS EXPIRY DATE: 10/16/2016
VIOLATION CODE: DUI
ARREST TIME: 21:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:55

SUBJECT'S LAST NAME: PINEDA FLORES

DATE/TIME OF TEST: 01/23/2016 22:23:46

Diagnostics Check		OK	22:15
Air Blank		0.000	22:15
Control Test		0.080	22:16
Air Blank		0.000	22:16
Subject Sample	# 1	0.194	22:18
Breath Volume		1.535	22:18
Air Blank		0.000	22:18
Air Blank		0.000	22:20
Subject Sample	# 2	0.200	22:22
Breath Volume		1.578	22:22
Air Blank		0.000	22:22
Control Test		0.079	22:23
Air Blank		0.000	22:23
Diagnostics Check		OK	22:23

Data Download Date/Time: 02/08/2016 15:24

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: APOPKA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPBELL,ROBERT,J
BREATH TEST OPERATOR AGENCY: Apopka PD
ARRESTING OFFICER: CAMPBELL,ROBERT
ARRESTING OFFICER AGENCY: Apopka PD
LAST AGENCY INSPECTION: 12/31/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: MERCER

DATE/TIME OF TEST: 01/01/2016 02:49:04

Diagnostics Check		OK	02:42
Air Blank		0.000	02:43
Control Test		0.077	02:43
Air Blank		0.000	02:44
Subject Sample	# 1	0.131	02:44
Breath Volume		2.214	02:44
Air Blank		0.000	02:45
Air Blank		0.000	02:47
Subject Sample	# 2	0.134	02:47
Breath Volume		2.312	02:47
Air Blank		0.000	02:48
Control Test		0.077	02:48
Air Blank		0.000	02:48
Diagnostics Check		OK	02:49

Data Download Date/Time: 01/31/2016 15:26

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: APOPKA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPBELL,ROBERT,J
BREATH TEST OPERATOR AGENCY: Apopka PD
ARRESTING OFFICER: CAMPBELL,ROBERT
ARRESTING OFFICER AGENCY: Apopka PD
LAST AGENCY INSPECTION: 12/31/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: VIVANCO

DATE/TIME OF TEST: 01/01/2016 03:34:56

Diagnostics Check		OK	03:28
Air Blank		0.000	03:28
Control Test		0.077	03:29
Air Blank		0.000	03:29
Subject Sample	# 1	0.000	03:30
Breath Volume		1.402	03:30
Air Blank		0.000	03:30
Air Blank		0.000	03:33
Subject Sample	# 2	0.000	03:33
Breath Volume		1.707	03:33
Air Blank		0.000	03:34
Control Test		0.077	03:34
Air Blank		0.000	03:34
Diagnostics Check		OK	03:34

Data Download Date/Time: 01/31/2016 15:26

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: APOPKA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPBELL,ROBERT,J
BREATH TEST OPERATOR AGENCY: Apopka PD
ARRESTING OFFICER: MELLON,JENNIFER
ARRESTING OFFICER AGENCY: Apopka PD
LAST AGENCY INSPECTION: 12/31/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: MCCOOK

DATE/TIME OF TEST: 01/06/2016 03:53:09

Diagnostics Check		OK	03:49	
Air Blank		0.000	03:49	
Control Test		0.077	03:49	
Air Blank		0.000	03:50	
Subject Sample	# 1		03:51	Subject Test Refused
Breath Volume		0.000	03:51	
Air Blank		0.000	03:52	
Control Test		0.077	03:52	
Air Blank		0.000	03:53	
Diagnostics Check		OK	03:53	

Data Download Date/Time: 01/31/2016 15:27

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: APOPKA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLER,ASHLEY,B
BREATH TEST OPERATOR AGENCY: Apopka PD
ARRESTING OFFICER: ELLER,ASHLEY
ARRESTING OFFICER AGENCY: Apopka PD
LAST AGENCY INSPECTION: 12/31/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:14

SUBJECT'S LAST NAME: DOMINGO CARRILLO

DATE/TIME OF TEST: 01/11/2016 02:46:37

Diagnostics Check		OK	02:40
Air Blank		0.000	02:40
Control Test		0.077	02:41
Air Blank		0.000	02:41
Subject Sample	# 1	0.179	02:42
Breath Volume		2.550	02:42
Air Blank		0.000	02:42
Air Blank		0.000	02:44
Subject Sample	# 2	0.190	02:45
Breath Volume		2.296	02:45
Air Blank		0.000	02:45
Control Test		0.077	02:46
Air Blank		0.000	02:46
Diagnostics Check		OK	02:46

Data Download Date/Time: 01/31/2016 15:27

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: APOPKA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPBELL,ROBERT,J
BREATH TEST OPERATOR AGENCY: Apopka PD
ARRESTING OFFICER: CAMPBELL,ROBERT
ARRESTING OFFICER AGENCY: Apopka PD
LAST AGENCY INSPECTION: 12/31/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: HERRERA

DATE/TIME OF TEST: 01/25/2016 01:27:29

Diagnostics Check		OK	01:20
Air Blank		0.000	01:20
Control Test		0.079	01:20
Air Blank		0.000	01:21
Subject Sample	# 1	0.141	01:22
Breath Volume		1.605	01:22
Air Blank		0.000	01:23
Air Blank		0.000	01:25
Subject Sample	# 2	0.143	01:25
Breath Volume		1.628	01:25
Air Blank		0.000	01:26
Control Test		0.078	01:26
Air Blank		0.000	01:27
Diagnostics Check		OK	01:27

Data Download Date/Time: 01/31/2016 15:27

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: APOPKA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPBELL,ROBERT,J
BREATH TEST OPERATOR AGENCY: Apopka PD
ARRESTING OFFICER: CAMPBELL,ROBERT
ARRESTING OFFICER AGENCY: Apopka PD
LAST AGENCY INSPECTION: 12/31/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:12

SUBJECT'S LAST NAME: GILLYARD

DATE/TIME OF TEST: 01/30/2016 23:38:09

Diagnostics Check		OK	23:35	
Air Blank		0.000	23:35	
Control Test		0.078	23:35	
Air Blank		0.000	23:36	
Subject Sample	# 1		23:36	Subject Test Refused
Breath Volume		0.000	23:36	
Air Blank		0.000	23:37	
Control Test		0.078	23:37	
Air Blank		0.000	23:38	
Diagnostics Check		OK	23:38	

Data Download Date/Time: 01/31/2016 15:27

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVA,EMILIANO,R
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: SILVA,EMILIANO
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 01/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1578837
DRY GAS EXPIRY DATE: 10/16/2016
VIOLATION CODE: DUI
ARREST TIME: 19:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:10

SUBJECT'S LAST NAME: FALASCO

DATE/TIME OF TEST: 01/08/2016 21:38:02

Diagnostics Check		OK	21:31
Air Blank		0.000	21:32
Control Test		0.081	21:32
Air Blank		0.000	21:32
Subject Sample	# 1	0.222	21:33
Breath Volume		3.031	21:33
Air Blank		0.000	21:34
Air Blank		0.000	21:36
Subject Sample	# 2	0.227	21:36
Breath Volume		2.472	21:36
Air Blank		0.000	21:37
Control Test		0.081	21:37
Air Blank		0.000	21:37
Diagnostics Check		OK	21:38

Data Download Date/Time: 02/08/2016 15:18

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUGO,NEFTALI,F
BREATH TEST OPERATOR AGENCY: Kissimmee PD
ARRESTING OFFICER: LUGO,NEFTALI
ARRESTING OFFICER AGENCY: Kissimmee PD
LAST AGENCY INSPECTION: 01/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1578837
DRY GAS EXPIRY DATE: 10/16/2016
VIOLATION CODE: DUI
ARREST TIME: 21:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:43

SUBJECT'S LAST NAME: GONZALES

DATE/TIME OF TEST: 01/08/2016 22:18:39

Diagnostics Check		OK	22:03	
Air Blank		0.000	22:04	
Control Test		0.081	22:04	
Air Blank		0.000	22:05	
Subject Sample	# 1	0.189	22:05	
Breath Volume		1.371	22:05	
Air Blank		0.000	22:06	
Air Blank		0.000	22:08	
Subject Sample	# 2	0.174	22:11	Slope Not Level
Breath Volume		1.250	22:11	
Air Blank		0.000	22:12	
Air Blank		0.000	22:13	
Subject Sample	# 3	0.179	22:17	Volume Not Met
Breath Volume		0.976	22:17	
Air Blank		0.000	22:17	
Control Test		0.080	22:18	
Air Blank		0.000	22:18	
Diagnostics Check		OK	22:18	

Data Download Date/Time: 02/08/2016 15:18

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUGO,NEFTALI,F
BREATH TEST OPERATOR AGENCY: Kissimmee PD
ARRESTING OFFICER: LUGO,NEFTALI
ARRESTING OFFICER AGENCY: Kissimmee PD
LAST AGENCY INSPECTION: 01/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1578837
DRY GAS EXPIRY DATE: 10/16/2016
VIOLATION CODE: DUI
ARREST TIME: 21:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:43

SUBJECT'S LAST NAME: GONZALES

DATE/TIME OF TEST: 01/08/2016 22:31:32

Diagnostics Check		OK	22:20
Air Blank		0.000	22:21
Control Test		0.080	22:21
Air Blank		0.000	22:21
Subject Sample	# 1	0.168	22:24
Breath Volume		1.222	22:24
Air Blank		0.000	22:25
Air Blank		0.000	22:26
Subject Sample	# 2	0.179	22:29
Breath Volume		1.539	22:29
Air Blank		0.000	22:30
Control Test		0.080	22:30
Air Blank		0.000	22:31
Diagnostics Check		OK	22:31

Data Download Date/Time: 02/08/2016 15:18

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JORGE,JUNIOR,
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: JORGE,JUNIOR
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 01/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1578837
DRY GAS EXPIRY DATE: 10/16/2016
VIOLATION CODE: DUI
ARREST TIME: 05:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:08

SUBJECT'S LAST NAME: ELALAMI

DATE/TIME OF TEST: 01/09/2016 06:40:06

Diagnostics Check		OK	06:33
Air Blank		0.000	06:34
Control Test		0.081	06:34
Air Blank		0.000	06:35
Subject Sample	# 1	0.042	06:35
Breath Volume		3.246	06:35
Air Blank		0.000	06:36
Air Blank		0.000	06:38
Subject Sample	# 2	0.040	06:38
Breath Volume		3.621	06:38
Air Blank		0.000	06:39
Control Test		0.082	06:39
Air Blank		0.000	06:40
Diagnostics Check		OK	06:40

Data Download Date/Time: 02/08/2016 15:18

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CABOT,GEORGE,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: PYE,CHRISTOPHER
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 01/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1578837
DRY GAS EXPIRY DATE: 10/16/2016
VIOLATION CODE: DUI
ARREST TIME: 06:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:10

SUBJECT'S LAST NAME: IGUINA MARTINEZ

DATE/TIME OF TEST: 01/10/2016 07:59:22

Diagnostics Check		OK	07:53
Air Blank		0.000	07:53
Control Test		0.081	07:54
Air Blank		0.000	07:54
Subject Sample	# 1	0.093	07:54
Breath Volume		1.472	07:54
Air Blank		0.000	07:55
Air Blank		0.000	07:57
Subject Sample	# 2	0.092	07:57
Breath Volume		2.179	07:57
Air Blank		0.000	07:58
Control Test		0.080	07:58
Air Blank		0.000	07:59
Diagnostics Check		OK	07:59

Data Download Date/Time: 02/08/2016 15:18

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONDELLO,ANTHONY,J
BREATH TEST OPERATOR AGENCY: St. Cloud PD
ARRESTING OFFICER: MONDELLO,ANTHONY
ARRESTING OFFICER AGENCY: St. Cloud PD
LAST AGENCY INSPECTION: 01/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1578837
DRY GAS EXPIRY DATE: 10/16/2016
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: MUSSO

DATE/TIME OF TEST: 01/12/2016 01:19:28

Diagnostics Check		OK	01:12
Air Blank		0.000	01:12
Control Test		0.082	01:13
Air Blank		0.000	01:13
Subject Sample	# 1	0.186	01:14
Breath Volume		3.292	01:14
Air Blank		0.000	01:15
Air Blank		0.000	01:17
Subject Sample	# 2	0.188	01:17
Breath Volume		3.464	01:17
Air Blank		0.000	01:18
Control Test		0.080	01:18
Air Blank		0.000	01:19
Diagnostics Check		OK	01:19

Data Download Date/Time: 02/08/2016 15:18

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUGO,NEFTALI,F
BREATH TEST OPERATOR AGENCY: Kissimmee PD
ARRESTING OFFICER: FORNEY,DANIEL
ARRESTING OFFICER AGENCY: Kissimmee PD
LAST AGENCY INSPECTION: 01/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1578837
DRY GAS EXPIRY DATE: 10/16/2016
VIOLATION CODE: DUI
ARREST TIME: 02:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: MOORE

DATE/TIME OF TEST: 01/14/2016 02:54:39

Diagnostics Check		OK	02:51	
Air Blank		0.000	02:52	
Control Test		0.081	02:52	
Air Blank		0.000	02:52	
Subject Sample	# 1		02:53	Subject Test Refused
Breath Volume		0.000	02:53	
Air Blank		0.000	02:53	
Control Test		0.081	02:54	
Air Blank		0.000	02:54	
Diagnostics Check		OK	02:54	

Data Download Date/Time: 02/08/2016 15:18

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLOW,CHRISTOPHER,A
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: ESH,J
ARRESTING OFFICER AGENCY: St. Cloud PD
LAST AGENCY INSPECTION: 01/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1578837
DRY GAS EXPIRY DATE: 10/16/2016
VIOLATION CODE: DUI
ARREST TIME: 01:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:46

SUBJECT'S LAST NAME: POWLESS

DATE/TIME OF TEST: 01/15/2016 02:17:29

Diagnostics Check		OK	02:10
Air Blank		0.000	02:11
Control Test		0.082	02:11
Air Blank		0.000	02:12
Subject Sample	# 1	0.118	02:13
Breath Volume		3.199	02:13
Air Blank		0.000	02:13
Air Blank		0.000	02:15
Subject Sample	# 2	0.115	02:15
Breath Volume		1.914	02:15
Air Blank		0.000	02:16
Control Test		0.081	02:16
Air Blank		0.000	02:17
Diagnostics Check		OK	02:17

Data Download Date/Time: 02/08/2016 15:18

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BINET,MARLON,J
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: BINET,MARLON
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 01/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1578837
DRY GAS EXPIRY DATE: 10/16/2016
VIOLATION CODE: DUI
ARREST TIME: 01:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: COOPER

DATE/TIME OF TEST: 01/17/2016 02:09:45

Diagnostics Check		OK	02:01
Air Blank		0.000	02:01
Control Test		0.082	02:02
Air Blank		0.000	02:02
Subject Sample	# 1	0.131	02:04
Breath Volume		1.320	02:04
Air Blank		0.000	02:05
Air Blank		0.000	02:07
Subject Sample	# 2	0.123	02:08
Breath Volume		2.031	02:08
Air Blank		0.000	02:08
Control Test		0.080	02:09
Air Blank		0.000	02:09
Diagnostics Check		OK	02:09

Data Download Date/Time: 02/08/2016 15:18

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONDELLO,ANTHONY,J
BREATH TEST OPERATOR AGENCY: St. Cloud PD
ARRESTING OFFICER: MONDELLO,ANTHONY
ARRESTING OFFICER AGENCY: St. Cloud PD
LAST AGENCY INSPECTION: 01/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1578837
DRY GAS EXPIRY DATE: 10/16/2016
VIOLATION CODE: DUI
ARREST TIME: 00:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:42

SUBJECT'S LAST NAME: SANCHEZ SANTOS

DATE/TIME OF TEST: 01/18/2016 01:12:32

Diagnostics Check		OK	01:05
Air Blank		0.000	01:06
Control Test		0.081	01:06
Air Blank		0.000	01:07
Subject Sample	# 1	0.178	01:07
Breath Volume		4.308	01:07
Air Blank		0.000	01:08
Air Blank		0.000	01:10
Subject Sample	# 2	0.181	01:11
Breath Volume		3.882	01:11
Air Blank		0.000	01:11
Control Test		0.080	01:11
Air Blank		0.000	01:12
Diagnostics Check		OK	01:12

Data Download Date/Time: 02/08/2016 15:18

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TELLERIA,TROY,R
BREATH TEST OPERATOR AGENCY: St. Cloud PD
ARRESTING OFFICER: SOUZA,WAYNE
ARRESTING OFFICER AGENCY: St. Cloud PD
LAST AGENCY INSPECTION: 01/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1578837
DRY GAS EXPIRY DATE: 10/16/2016
VIOLATION CODE: DUI
ARREST TIME: 01:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:17

SUBJECT'S LAST NAME: DESKINS

DATE/TIME OF TEST: 01/18/2016 02:44:32

Diagnostics Check		OK	02:37
Air Blank		0.000	02:38
Control Test		0.081	02:38
Air Blank		0.000	02:39
Subject Sample	# 1	0.000	02:40
Breath Volume		1.933	02:40
Air Blank		0.000	02:40
Air Blank		0.000	02:42
Subject Sample	# 2	0.000	02:43
Breath Volume		2.085	02:43
Air Blank		0.000	02:43
Control Test		0.081	02:43
Air Blank		0.000	02:44
Diagnostics Check		OK	02:44

Data Download Date/Time: 02/08/2016 15:18

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPOS,MICHAEL,A
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: BEEKMAN,JACOB
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 01/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1578837
DRY GAS EXPIRY DATE: 10/16/2016
VIOLATION CODE: DUI
ARREST TIME: 03:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:34

SUBJECT'S LAST NAME: SOTO LOPEZ

DATE/TIME OF TEST: 01/18/2016 04:00:51

Diagnostics Check		OK	03:54
Air Blank		0.000	03:55
Control Test		0.082	03:55
Air Blank		0.000	03:55
Subject Sample	# 1	0.179	03:56
Breath Volume		1.582	03:56
Air Blank		0.000	03:57
Air Blank		0.000	03:58
Subject Sample	# 2	0.177	03:59
Breath Volume		1.812	03:59
Air Blank		0.000	03:59
Control Test		0.081	04:00
Air Blank		0.000	04:00
Diagnostics Check		OK	04:00

Data Download Date/Time: 02/08/2016 15:18

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRAWFORD,DAVID,T
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: CRAWFORD,DAVID
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 01/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1578837
DRY GAS EXPIRY DATE: 10/16/2016
VIOLATION CODE: DUI
ARREST TIME: 03:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:29

SUBJECT'S LAST NAME: LUQUE LOPEZ

DATE/TIME OF TEST: 01/18/2016 04:05:47

Diagnostics Check	OK	04:05	
Air Blank		04:05	RFI Detect
Air Blank	0.000	04:05	

Data Download Date/Time: 02/08/2016 15:19

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRAWFORD,DAVID,T
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: CRAWFORD,DAVID
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 01/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1578837
DRY GAS EXPIRY DATE: 10/16/2016
VIOLATION CODE: DUI
ARREST TIME: 03:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:29

SUBJECT'S LAST NAME: LUQUE LOPEZ

DATE/TIME OF TEST: 01/18/2016 04:13:52

Diagnostics Check		OK	04:07
Air Blank		0.000	04:08
Control Test		0.082	04:08
Air Blank		0.000	04:08
Subject Sample	# 1	0.184	04:09
Breath Volume		1.898	04:09
Air Blank		0.000	04:10
Air Blank		0.000	04:11
Subject Sample	# 2	0.181	04:12
Breath Volume		2.425	04:12
Air Blank		0.000	04:12
Control Test		0.078	04:13
Air Blank		0.000	04:13
Diagnostics Check		OK	04:13

Data Download Date/Time: 02/08/2016 15:19

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALTHER,PAUL,M
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: PYE,CHRISTOPHER
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 01/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1578837
DRY GAS EXPIRY DATE: 10/16/2016
VIOLATION CODE: DUI
ARREST TIME: 21:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:29

SUBJECT'S LAST NAME: VARGAS ALEMAN

DATE/TIME OF TEST: 01/18/2016 22:08:34

Diagnostics Check		OK	22:02
Air Blank		0.000	22:03
Control Test		0.081	22:03
Air Blank		0.000	22:03
Subject Sample	# 1	0.000	22:04
Breath Volume		1.406	22:04
Air Blank		0.000	22:04
Air Blank		0.000	22:06
Subject Sample	# 2	0.000	22:07
Breath Volume		2.898	22:07
Air Blank		0.000	22:07
Control Test		0.082	22:07
Air Blank		0.000	22:08
Diagnostics Check		OK	22:08

Data Download Date/Time: 02/08/2016 15:19

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALTHER,PAUL,M
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: PYE,CHRISTOPHER
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 01/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1578837
DRY GAS EXPIRY DATE: 10/16/2016
VIOLATION CODE: DUI
ARREST TIME: 03:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:18

SUBJECT'S LAST NAME: TAVAREZ

DATE/TIME OF TEST: 01/20/2016 04:04:44

Diagnostics Check		OK	03:58
Air Blank		0.000	03:58
Control Test		0.082	03:59
Air Blank		0.000	03:59
Subject Sample	# 1	0.139	04:00
Breath Volume		2.421	04:00
Air Blank		0.000	04:00
Air Blank		0.000	04:02
Subject Sample	# 2	0.140	04:03
Breath Volume		2.683	04:03
Air Blank		0.000	04:03
Control Test		0.082	04:04
Air Blank		0.000	04:04
Diagnostics Check		OK	04:04

Data Download Date/Time: 02/08/2016 15:19

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALTHER,PAUL,M
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: PYE,CHRISTOPHER
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 01/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1578837
DRY GAS EXPIRY DATE: 10/16/2016
VIOLATION CODE: DUI
ARREST TIME: 01:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: WATSON

DATE/TIME OF TEST: 01/23/2016 02:11:47

Diagnostics Check		OK	02:05
Air Blank		0.000	02:06
Control Test		0.081	02:06
Air Blank		0.000	02:06
Subject Sample	# 1	0.068	02:07
Breath Volume		2.742	02:07
Air Blank		0.000	02:07
Air Blank		0.000	02:09
Subject Sample	# 2	0.069	02:10
Breath Volume		2.785	02:10
Air Blank		0.000	02:10
Control Test		0.081	02:11
Air Blank		0.000	02:11
Diagnostics Check		OK	02:11

Data Download Date/Time: 02/08/2016 15:19

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALTHER,PAUL,M
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: CORUJO,PEDRO
ARRESTING OFFICER AGENCY: Kissimmee PD
LAST AGENCY INSPECTION: 01/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1578837
DRY GAS EXPIRY DATE: 10/16/2016
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: NUNEZ IRIZARRY

DATE/TIME OF TEST: 01/23/2016 02:40:27

Diagnostics Check		OK	02:30
Air Blank		0.000	02:31
Control Test		0.081	02:31
Air Blank		0.000	02:32
Subject Sample	# 1	0.227	02:34
Breath Volume		1.730	02:34
Air Blank		0.000	02:34
Air Blank		0.000	02:36
Subject Sample	# 2	0.218	02:38
Breath Volume		1.761	02:38
Air Blank		0.000	02:39
Control Test		0.081	02:39
Air Blank		0.000	02:40
Diagnostics Check		OK	02:40

Data Download Date/Time: 02/08/2016 15:19

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARI,ADAM,
BREATH TEST OPERATOR AGENCY: St. Cloud PD
ARRESTING OFFICER: ESH,JOSH
ARRESTING OFFICER AGENCY: St. Cloud PD
LAST AGENCY INSPECTION: 01/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1578837
DRY GAS EXPIRY DATE: 10/16/2016
VIOLATION CODE: DUI
ARREST TIME: 03:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: HAIRSTON

DATE/TIME OF TEST: 01/24/2016 04:28:18

Diagnostics Check		OK	04:22
Air Blank		0.000	04:22
Control Test		0.080	04:23
Air Blank		0.000	04:23
Subject Sample	# 1	0.080	04:23
Breath Volume		1.332	04:23
Air Blank		0.000	04:24
Air Blank		0.000	04:26
Subject Sample	# 2	0.077	04:26
Breath Volume		1.738	04:26
Air Blank		0.000	04:27
Control Test		0.081	04:27
Air Blank		0.000	04:28
Diagnostics Check		OK	04:28

Data Download Date/Time: 02/08/2016 15:19

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONDELLO,ANTHONY,J
BREATH TEST OPERATOR AGENCY: St. Cloud PD
ARRESTING OFFICER: MONDELLO,ANTHONY
ARRESTING OFFICER AGENCY: St. Cloud PD
LAST AGENCY INSPECTION: 01/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1578837
DRY GAS EXPIRY DATE: 10/16/2016
VIOLATION CODE: DUI
ARREST TIME: 00:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:36

SUBJECT'S LAST NAME: GERSTENLAUER

DATE/TIME OF TEST: 01/30/2016 01:14:27

Diagnostics Check		OK	00:59	
Air Blank		0.000	00:59	
Control Test		0.081	01:00	
Air Blank		0.000	01:00	
Subject Sample	# 1	0.184	01:03	Volume Not Met
Breath Volume		0.097	01:03	
Air Blank		0.000	01:04	
Air Blank		0.000	01:06	
Subject Sample	# 2	0.144	01:07	
Breath Volume		1.812	01:07	
Air Blank		0.000	01:07	
Air Blank		0.000	01:09	
Subject Sample	# 3	0.170	01:12	Volume Not Met
Breath Volume		0.781	01:12	
Air Blank		0.000	01:13	
Control Test		0.080	01:13	
Air Blank		0.000	01:14	
Diagnostics Check		OK	01:14	

Data Download Date/Time: 02/08/2016 15:19

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BEEKMAN,JACOB,C
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: PEREZ,JUAN
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 01/05/2016

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1578837
DRY GAS EXPIRY DATE: 10/16/2016
VIOLATION CODE: DUI
ARREST TIME: 19:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:25

SUBJECT'S LAST NAME: GLEMBIN

DATE/TIME OF TEST: 01/31/2016 20:53:31

Diagnostics Check		OK	20:46
Air Blank		0.000	20:46
Control Test		0.079	20:46
Air Blank		0.000	20:47
Subject Sample	# 1	0.266	20:48
Breath Volume		1.562	20:48
Air Blank		0.000	20:49
Air Blank		0.000	20:51
Subject Sample	# 2	0.269	20:51
Breath Volume		1.324	20:51
Air Blank		0.000	20:52
Control Test		0.079	20:52
Air Blank		0.000	20:53
Diagnostics Check		OK	20:53

Data Download Date/Time: 02/08/2016 15:19

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VERSAGGI,JOSEPH,
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: MEEHAN,RYAN
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 12/15/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 36314080A3
DRY GAS EXPIRY DATE: 02/05/2017
VIOLATION CODE: DUI
ARREST TIME: 19:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:35

SUBJECT'S LAST NAME: SOTO

DATE/TIME OF TEST: 01/10/2016 20:10:59

Diagnostics Check		OK	20:04
Air Blank		0.000	20:04
Control Test		0.080	20:05
Air Blank		0.000	20:05
Subject Sample	# 1	0.189	20:06
Breath Volume		3.066	20:06
Air Blank		0.000	20:07
Air Blank		0.000	20:08
Subject Sample	# 2	0.190	20:09
Breath Volume		3.257	20:09
Air Blank		0.000	20:10
Control Test		0.080	20:10
Air Blank		0.000	20:10
Diagnostics Check		OK	20:10

Data Download Date/Time: 01/30/2016 16:13

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.