Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: PONDER,CHRISTOPHER
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 01/27/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 00:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: SEPULUEDA-TORRES

DATE/TIME OF TEST: 01/29/2016 01:14:12

Diagnostics Check		OK	01:07
Air Blank		0.000	01:08
Control Test		0.079	01:08
Air Blank		0.000	01:09
Subject Sample	# 1	0.214	01:09
Breath Volume		2.593	01:09
Air Blank		0.000	01:10
Air Blank		0.000	01:12
Subject Sample	# 2	0.221	01:12
Breath Volume		3.261	01:12
Air Blank		0.000	01:13
Control Test		0.079	01:13
Air Blank		0.000	01:14
Diagnostics Check		OK	01:14

Data Download Date/Time: 01/29/2016 20:28

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: COLBURN,MEGAN
ARRESTING OFFICER AGENCY: Casselberry PD
LAST AGENCY INSPECTION: 01/27/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 01:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:11

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 01/29/2016 02:38:55

Diagnostics Check		OK	02:32
Air Blank		0.000	02:32
Control Test		0.079	02:33
Air Blank		0.000	02:33
Subject Sample	# 1	0.124	02:34
Breath Volume		3.050	02:34
Air Blank		0.000	02:34
Air Blank		0.000	02:36
Subject Sample	# 2	0.124	02:37
Breath Volume		3.019	02:37
Air Blank		0.000	02:38
Control Test		0.080	02:38
Air Blank		0.000	02:38
Diagnostics Check		OK	02:38

Data Download Date/Time: 01/29/2016 20:28

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: FLANAGAN,BRENT
ARRESTING OFFICER AGENCY: Sanford PD
LAST AGENCY INSPECTION: 01/27/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 02:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:34

SUBJECT'S LAST NAME: BORTHWICK

DATE/TIME OF TEST: 01/29/2016 04:08:05

Diagnostics Check		OK	03:55	
Air Blank		0.000	03:56	
Control Test		0.079	03:56	
Air Blank		0.000	03:57	
Subject Sample	# 1	0.085	04:00	Volume Not Met
Breath Volume		0.292	04:00	
Air Blank		0.000	04:00	
Air Blank		0.000	04:02	
Subject Sample	# 2	0.130	04:03	
Breath Volume		2.800	04:03	
Air Blank		0.000	04:03	
Air Blank		0.000	04:05	
Subject Sample	# 3	0.118	04:06	
Breath Volume		1.398	04:06	
Air Blank		0.000	04:07	
Control Test		0.081	04:07	
Air Blank		0.000	04:07	
Diagnostics Check		OK	04:08	

Data Download Date/Time: 01/29/2016 20:28

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: BARNHORST,EDDIE
ARRESTING OFFICER AGENCY: Winter Springs PD
LAST AGENCY INSPECTION: 01/27/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 03:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:06

SUBJECT'S LAST NAME: BANDA

DATE/TIME OF TEST: 01/29/2016 04:32:51

Diagnostics Check		OK	04:26
Air Blank		0.000	04:27
Control Test		0.080	04:27
Air Blank		0.000	04:27
Subject Sample	# 1	0.142	04:28
Breath Volume		1.898	04:28
Air Blank		0.000	04:28
Air Blank		0.000	04:30
Subject Sample	# 2	0.155	04:31
Breath Volume		2.984	04:31
Air Blank		0.000	04:31
Control Test		0.080	04:32
Air Blank		0.000	04:32
Diagnostics Check		OK	04:32

Data Download Date/Time: 01/29/2016 20:28

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000999

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: FOWLER,MATTHEW
ARRESTING OFFICER AGENCY: Altamonte Springs PD
LAST AGENCY INSPECTION: 01/27/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1787697
DRY GAS EXPIRY DATE: 01/08/2018
VIOLATION CODE: DUI
ARREST TIME: 01:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:28

SUBJECT'S LAST NAME: RILEY

DATE/TIME OF TEST: 01/28/2016 01:55:21

Diagnostics Check		OK	01:48
Air Blank		0.000	01:48
Control Test		0.078	01:49
Air Blank		0.000	01:49
Subject Sample	# 1	0.218	01:50
Breath Volume		3.121	01:50
Air Blank		0.000	01:51
Air Blank		0.000	01:52
Subject Sample	# 2	0.204	01:53
Breath Volume		2.167	01:53
Air Blank		0.000	01:54
Control Test		0.078	01:54
Air Blank		0.000	01:55
Diagnostics Check		OK	01:55

Data Download Date/Time: 01/29/2016 20:35

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000999

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: BRYANT,STEVEN
ARRESTING OFFICER AGENCY: Longwood PD
LAST AGENCY INSPECTION: 01/27/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1787697
DRY GAS EXPIRY DATE: 01/08/2018
VIOLATION CODE: DUI
ARREST TIME: 01:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:53

SUBJECT'S LAST NAME: ROSALIA

DATE/TIME OF TEST: 01/28/2016 03:27:24

Diagnostics Check		OK	03:20
Air Blank		0.000	03:20
Control Test		0.077	03:20
Air Blank		0.000	03:21
Subject Sample	# 1	0.195	03:23
Breath Volume		1.625	03:23
Air Blank		0.000	03:23
Air Blank		0.000	03:25
Subject Sample	# 2	0.202	03:25
Breath Volume		1.453	03:25
Air Blank		0.000	03:26
Control Test		0.077	03:26
Air Blank		0.000	03:27
Diagnostics Check		OK	03:27

Data Download Date/Time: 01/29/2016 20:35

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001000

INSTRUMENT REGISTERED TO: COCOA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAPTONSTALL,BRENT,C
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: BIVINS,HEATHER
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 12/23/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A2
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:06

SUBJECT'S LAST NAME: WEST

DATE/TIME OF TEST: 01/11/2016 03:27:07

Diagnostics Check		OK	03:20
Air Blank		0.000	03:21
Control Test		0.081	03:21
Air Blank		0.000	03:22
Subject Sample	# 1	0.105	03:22
Breath Volume		2.625	03:22
Air Blank		0.000	03:23
Air Blank		0.000	03:25
Subject Sample	# 2	0.103	03:25
Breath Volume		2.011	03:25
Air Blank		0.000	03:26
Control Test		0.080	03:26
Air Blank		0.000	03:27
Diagnostics Check		OK	03:27

Data Download Date/Time: 01/30/2016 21:27

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRIEBELL,DON,R
BREATH TEST OPERATOR AGENCY: Satellite Beach PD
ARRESTING OFFICER: HART,BRIAN
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 12/27/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1878845
DRY GAS EXPIRY DATE: 06/22/2018
VIOLATION CODE: DUI
ARREST TIME: 01:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:52

SUBJECT'S LAST NAME: GORDON

DATE/TIME OF TEST: 01/03/2016 04:01:50

Diagnostics Check		OK	03:51	
Air Blank		0.000	03:51	
Control Test		0.080	03:52	
Air Blank		0.000	03:52	
Subject Sample	# 1	0.081	03:55	Volume Not Met
Breath Volume		0.125	03:55	
Air Blank		0.000	03:56	
Air Blank		0.000	03:58	
Subject Sample	# 2		04:00	Subject Test Refused
Breath Volume		0.000	04:00	
Air Blank		0.000	04:00	
Control Test		0.078	04:01	
Air Blank		0.000	04:01	
Diagnostics Check		OK	04:01	

Data Download Date/Time: 01/29/2016 20:37

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRIEBELL,DON,R
BREATH TEST OPERATOR AGENCY: Satellite Beach PD
ARRESTING OFFICER: DUMMER,LUKE
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 12/27/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1878845
DRY GAS EXPIRY DATE: 06/22/2018
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: L344110837200

DATE/TIME OF TEST: 01/03/2016 04:13:13

Diagnostics Check		OK	04:06
Air Blank		0.000	04:06
Control Test		0.079	04:06
Air Blank		0.000	04:07
Subject Sample	# 1	0.164	04:08
Breath Volume		1.812	04:08
Air Blank		0.000	04:08
Air Blank		0.000	04:10
Subject Sample	# 2	0.172	04:11
Breath Volume		2.074	04:11
Air Blank		0.000	04:12
Control Test		0.079	04:12
Air Blank		0.000	04:13
Diagnostics Check		OK	04:13

Data Download Date/Time: 01/29/2016 20:37

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRIEBELL,DON,R
BREATH TEST OPERATOR AGENCY: Satellite Beach PD
ARRESTING OFFICER: HART,BRIAN
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 12/27/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1878845
DRY GAS EXPIRY DATE: 06/22/2018
VIOLATION CODE: DUI
ARREST TIME: 01:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:58

SUBJECT'S LAST NAME: GORDON

DATE/TIME OF TEST: 01/03/2016 04:28:50

Diagnostics Check		OK	04:22
Air Blank		0.000	04:22
Control Test		0.079	04:23
Air Blank		0.000	04:23
Subject Sample	# 1	0.095	04:24
Breath Volume		1.316	04:24
Air Blank		0.000	04:24
Air Blank		0.000	04:26
Subject Sample	# 2	0.090	04:27
Breath Volume		1.242	04:27
Air Blank		0.000	04:27
Control Test		0.080	04:28
Air Blank		0.000	04:28
Diagnostics Check		OK	04:28

Data Download Date/Time: 01/29/2016 20:37

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRIEBELL,DON,R
BREATH TEST OPERATOR AGENCY: Satellite Beach PD
ARRESTING OFFICER: TRIEBELL,DON
ARRESTING OFFICER AGENCY: Satellite Beach PD
LAST AGENCY INSPECTION: 12/27/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1878845
DRY GAS EXPIRY DATE: 06/22/2018
VIOLATION CODE: DUI
ARREST TIME: 00:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:01

SUBJECT'S LAST NAME: CALDERON ORTIZ

DATE/TIME OF TEST: 01/04/2016 01:37:10

Diagnostics Check		OK	01:30
Air Blank		0.000	01:30
Control Test		0.079	01:30
Air Blank		0.000	01:31
Subject Sample	# 1	0.176	01:32
Breath Volume		1.953	01:32
Air Blank		0.000	01:33
Air Blank		0.000	01:35
Subject Sample	# 2	0.188	01:35
Breath Volume		2.949	01:35
Air Blank		0.000	01:36
Control Test		0.079	01:36
Air Blank		0.000	01:37
Diagnostics Check		OK	01:37

Data Download Date/Time: 01/29/2016 20:37

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,ANTHONY,M
BREATH TEST OPERATOR AGENCY: Satellite Beach PD
ARRESTING OFFICER: HOLLAND,ANTHONY
ARRESTING OFFICER AGENCY: Satellite Beach PD
LAST AGENCY INSPECTION: 12/27/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1878845
DRY GAS EXPIRY DATE: 06/22/2018
VIOLATION CODE: DUI
ARREST TIME: 02:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:11

SUBJECT'S LAST NAME: COMER

DATE/TIME OF TEST: 01/10/2016 03:39:35

Diagnostics Check		OK	03:32
Air Blank		0.000	03:32
Control Test		0.079	03:32
Air Blank		0.000	03:33
Subject Sample	# 1	0.128	03:35
Breath Volume		1.281	03:35
Air Blank		0.000	03:35
Air Blank		0.000	03:37
Subject Sample	# 2	0.132	03:38
Breath Volume		1.574	03:38
Air Blank		0.000	03:38
Control Test		0.078	03:38
Air Blank		0.000	03:39
Diagnostics Check		OK	03:39

Data Download Date/Time: 01/29/2016 20:37

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRIEBELL,DON,R
BREATH TEST OPERATOR AGENCY: Satellite Beach PD
ARRESTING OFFICER: MATTHEWS,KRISTOPHER
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 12/27/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1878845
DRY GAS EXPIRY DATE: 06/22/2018
VIOLATION CODE: DUI
ARREST TIME: 04:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:15

SUBJECT'S LAST NAME: CARLISLE

DATE/TIME OF TEST: 01/13/2016 04:42:51

Diagnostics Check		OK	04:35
Air Blank		0.000	04:36
Control Test		0.080	04:36
Air Blank		0.000	04:37
Subject Sample	# 1	0.182	04:38
Breath Volume		1.546	04:38
Air Blank		0.000	04:38
Air Blank		0.000	04:40
Subject Sample	# 2	0.175	04:41
Breath Volume		2.535	04:41
Air Blank		0.000	04:41
Control Test		0.079	04:42
Air Blank		0.000	04:42
Diagnostics Check		OK	04:42

Data Download Date/Time: 01/29/2016 20:37

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROOKS,ROBBIE,M
BREATH TEST OPERATOR AGENCY: Satellite Beach PD
ARRESTING OFFICER: MCCRANN,SEAN
ARRESTING OFFICER AGENCY: Satellite Beach PD
LAST AGENCY INSPECTION: 12/27/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1878845
DRY GAS EXPIRY DATE: 06/22/2018
VIOLATION CODE: DUI
ARREST TIME: 02:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:56

SUBJECT'S LAST NAME: LEINBACK

DATE/TIME OF TEST: 01/16/2016 03:38:33

Diagnostics Check		OK	03:26	
Air Blank		0.000	03:27	
Control Test		0.079	03:27	
Air Blank		0.000	03:28	
Subject Sample	# 1	0.168	03:31	Volume Not Met
Breath Volume		0.015	03:31	
Air Blank		0.000	03:31	
Air Blank		0.000	03:33	
Subject Sample	# 2	0.177	03:36	Volume Not Met
Breath Volume		0.050	03:36	
Air Blank		0.000	03:37	
Control Test		0.077	03:37	
Air Blank		0.000	03:38	
Diagnostics Check		OK	03:38	

Data Download Date/Time: 01/29/2016 20:37

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROOKS,ROBBIE,M
BREATH TEST OPERATOR AGENCY: Satellite Beach PD
ARRESTING OFFICER: MCCRANN,SEAN
ARRESTING OFFICER AGENCY: Satellite Beach PD
LAST AGENCY INSPECTION: 12/27/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1878845
DRY GAS EXPIRY DATE: 06/22/2018
VIOLATION CODE: DUI
ARREST TIME: 02:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:56

SUBJECT'S LAST NAME: LEINBACK

DATE/TIME OF TEST: 01/16/2016 03:56:48

Diagnostics Check		OK	03:45	
Air Blank		0.000	03:45	
Control Test		0.078	03:46	
Air Blank		0.000	03:46	
Subject Sample	# 1	0.184	03:49	Volume Not Met
Breath Volume		0.050	03:49	
Air Blank		0.000	03:50	
Air Blank		0.000	03:52	
Subject Sample	# 2		03:55	No Sample Provided
Breath Volume		0.000	03:55	
Air Blank		0.000	03:55	
Control Test		0.079	03:56	
Air Blank		0.000	03:56	
Diagnostics Check		OK	03:56	

Data Download Date/Time: 01/29/2016 20:37

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRIEBELL,DON,R
BREATH TEST OPERATOR AGENCY: Satellite Beach PD
ARRESTING OFFICER: TRIEBELL,DON
ARRESTING OFFICER AGENCY: Satellite Beach PD
LAST AGENCY INSPECTION: 12/27/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1878845
DRY GAS EXPIRY DATE: 06/22/2018
VIOLATION CODE: DUI
ARREST TIME: 22:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:39

SUBJECT'S LAST NAME: JEWELL

DATE/TIME OF TEST: 01/20/2016 23:16:00

Diagnostics Check		OK	23:04	
Air Blank		0.000	23:04	
Control Test		0.080	23:05	
Air Blank		0.000	23:05	
Subject Sample	# 1	0.103	23:08	Volume Not Met
Breath Volume		0.253	23:08	
Air Blank		0.000	23:09	
Air Blank		0.000	23:11	
Subject Sample	# 2	0.110	23:14	Volume Not Met
Breath Volume		0.351	23:14	
Air Blank		0.000	23:15	
Control Test		0.078	23:15	
Air Blank		0.000	23:15	
Diagnostics Check		OK	23:16	

Data Download Date/Time: 01/29/2016 20:37

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SOUSA,PAUL,M
BREATH TEST OPERATOR AGENCY: Satellite Beach PD
ARRESTING OFFICER: SOUSA,PAUL
ARRESTING OFFICER AGENCY: Satellite Beach PD
LAST AGENCY INSPECTION: 12/27/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1878845
DRY GAS EXPIRY DATE: 06/22/2018
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:46

SUBJECT'S LAST NAME: MEARS

DATE/TIME OF TEST: 01/22/2016 02:20:19

Diagnostics Check		OK	02:13
Air Blank		0.000	02:14
Control Test		0.081	02:14
Air Blank		0.000	02:15
Subject Sample	# 1	0.185	02:15
Breath Volume		4.019	02:15
Air Blank		0.000	02:16
Air Blank		0.000	02:18
Subject Sample	# 2	0.184	02:18
Breath Volume		3.582	02:18
Air Blank		0.000	02:19
Control Test		0.080	02:19
Air Blank		0.000	02:20
Diagnostics Check		OK	02:20

Data Download Date/Time: 01/29/2016 20:37

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRIEBELL,DON,R
BREATH TEST OPERATOR AGENCY: Satellite Beach PD
ARRESTING OFFICER: TRIEBELL,DON
ARRESTING OFFICER AGENCY: Satellite Beach PD
LAST AGENCY INSPECTION: 12/27/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1878845
DRY GAS EXPIRY DATE: 06/22/2018
VIOLATION CODE: DUI
ARREST TIME: 02:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 01/23/2016 03:14:16

Diagnostics Check		OK	03:08
Air Blank		0.000	03:08
Control Test		0.080	03:08
Air Blank		0.000	03:09
Subject Sample	# 1	0.139	03:09
Breath Volume		1.378	03:09
Air Blank		0.000	03:10
Air Blank		0.000	03:12
Subject Sample	# 2	0.132	03:12
Breath Volume		1.445	03:12
Air Blank		0.000	03:13
Control Test		0.079	03:13
Air Blank		0.000	03:14
Diagnostics Check		OK	03:14

Data Download Date/Time: 01/29/2016 20:37

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROOKS,ROBBIE,M
BREATH TEST OPERATOR AGENCY: Satellite Beach PD
ARRESTING OFFICER: MCCRANN,SEAN
ARRESTING OFFICER AGENCY: Satellite Beach PD
LAST AGENCY INSPECTION: 12/27/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1878845
DRY GAS EXPIRY DATE: 06/22/2018
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:13

SUBJECT'S LAST NAME: LITTLE

DATE/TIME OF TEST: 01/25/2016 01:58:34

Diagnostics Check		OK	01:52
Air Blank		0.000	01:52
Control Test		0.080	01:52
Air Blank		0.000	01:53
Subject Sample	# 1	0.178	01:54
Breath Volume		2.023	01:54
Air Blank		0.000	01:54
Air Blank		0.000	01:56
Subject Sample	# 2	0.187	01:57
Breath Volume		2.140	01:57
Air Blank		0.000	01:57
Control Test		0.080	01:57
Air Blank		0.000	01:58
Diagnostics Check		OK	01:58

Data Download Date/Time: 01/29/2016 20:37

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PERSAUD,COMPTON,D
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KASELAK,STEPHEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 23:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:54

SUBJECT'S LAST NAME: TOLEDO

DATE/TIME OF TEST: 01/01/2016 00:24:01

Diagnostics Check		OK	00:17
Air Blank		0.000	00:18
Control Test		0.081	00:18
Air Blank		0.000	00:19
Subject Sample	# 1	0.180	00:19
Breath Volume		2.257	00:19
Air Blank		0.000	00:20
Air Blank		0.000	00:22
Subject Sample	# 2	0.170	00:22
Breath Volume		2.300	00:22
Air Blank		0.000	00:23
Control Test		0.081	00:23
Air Blank		0.000	00:23
Diagnostics Check		OK	00:24

Data Download Date/Time: 02/01/2016 07:03

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: CRIM,T
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 02:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: DALTON

DATE/TIME OF TEST: 01/01/2016 03:42:48

Diagnostics Check		OK	03:36
Air Blank		0.000	03:36
Control Test		0.081	03:36
Air Blank		0.000	03:37
Subject Sample	# 1	0.174	03:38
Breath Volume		2.714	03:38
Air Blank		0.000	03:38
Air Blank		0.000	03:40
Subject Sample	# 2	0.154	03:41
Breath Volume		2.894	03:41
Air Blank		0.000	03:41
Control Test		0.081	03:42
Air Blank		0.000	03:42
Diagnostics Check		OK	03:42

Data Download Date/Time: 02/01/2016 07:03

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ASHCOM,TIFFANY,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ASHCOM,TIFFANY
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 03:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: ZUNIGA

DATE/TIME OF TEST: 01/01/2016 04:21:51

Diagnostics Check		OK	04:15
Air Blank		0.000	04:16
Control Test		0.080	04:16
Air Blank		0.000	04:17
Subject Sample	# 1	0.150	04:17
Breath Volume		2.023	04:17
Air Blank		0.000	04:18
Air Blank		0.000	04:19
Subject Sample	# 2	0.143	04:20
Breath Volume		1.589	04:20
Air Blank		0.000	04:20
Control Test		0.080	04:21
Air Blank		0.000	04:21
Diagnostics Check		OK	04:21

Data Download Date/Time: 02/01/2016 07:03

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ASHCOM,TIFFANY,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ASHCOM,TIFFANY
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 03:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:44

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 01/02/2016 04:11:13

Diagnostics Check		OK	04:08	
Air Blank		0.000	04:08	
Control Test		0.081	04:09	
Air Blank		0.000	04:09	
Subject Sample	# 1		04:09	Subject Test Refused
Breath Volume		0.000	04:09	
Air Blank		0.000	04:10	
Control Test		0.081	04:10	
Air Blank		0.000	04:11	
Diagnostics Check		OK	04:11	

Data Download Date/Time: 02/01/2016 07:03

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PERSAUD,COMPTON,D
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: PERSAUD,COMPTON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 03:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:09

SUBJECT'S LAST NAME: HUFF

DATE/TIME OF TEST: 01/03/2016 03:36:21

Diagnostics Check		OK	03:29
Air Blank		0.000	03:30
Control Test		0.081	03:30
Air Blank		0.000	03:31
Subject Sample	# 1	0.169	03:31
Breath Volume		1.300	03:31
Air Blank		0.000	03:32
Air Blank		0.000	03:34
Subject Sample	# 2	0.178	03:34
Breath Volume		1.605	03:34
Air Blank		0.000	03:35
Control Test		0.081	03:35
Air Blank		0.000	03:36
Diagnostics Check		OK	03:36

Data Download Date/Time: 02/01/2016 07:03

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SWANNER,SARAH,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ROSADO,IAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 18:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:40

SUBJECT'S LAST NAME: CRISLER

DATE/TIME OF TEST: 01/03/2016 20:07:45

Diagnostics Check		OK	20:00
Air Blank		0.000	20:01
Control Test		0.081	20:01
Air Blank		0.000	20:01
Subject Sample	# 1	0.182	20:03
Breath Volume		2.363	20:03
Air Blank		0.000	20:03
Air Blank		0.000	20:05
Subject Sample	# 2	0.185	20:06
Breath Volume		3.300	20:06
Air Blank		0.000	20:06
Control Test		0.081	20:07
Air Blank		0.000	20:07
Diagnostics Check		OK	20:07

Data Download Date/Time: 02/01/2016 07:03

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ASHCOM,TIFFANY,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ASHCOM,TIFFANY
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 01:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:02

SUBJECT'S LAST NAME: HOPTRY

DATE/TIME OF TEST: 01/08/2016 02:28:27

Diagnostics Check		OK	02:22
Air Blank		0.000	02:23
Control Test		0.079	02:23
Air Blank		0.000	02:23
Subject Sample	# 1	0.145	02:24
Breath Volume		1.996	02:24
Air Blank		0.000	02:24
Air Blank		0.000	02:26
Subject Sample	# 2	0.150	02:26
Breath Volume		1.847	02:26
Air Blank		0.000	02:27
Control Test		0.080	02:27
Air Blank		0.000	02:28
Diagnostics Check		OK	02:28

Data Download Date/Time: 02/01/2016 07:03

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EASTTY,MARK,L
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: EASTTY,MARK
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 13:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:18

SUBJECT'S LAST NAME: HUDSON

DATE/TIME OF TEST: 01/08/2016 13:50:44

Diagnostics Check		OK	13:44
Air Blank		0.000	13:45
Control Test		0.081	13:45
Air Blank		0.000	13:46
Subject Sample	# 1	0.000	13:46
Breath Volume		1.652	13:46
Air Blank		0.000	13:47
Air Blank		0.000	13:49
Subject Sample	# 2	0.000	13:49
Breath Volume		1.984	13:49
Air Blank		0.000	13:49
Control Test		0.080	13:50
Air Blank		0.000	13:50
Diagnostics Check		OK	13:50

Data Download Date/Time: 02/01/2016 07:03

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ASHCOM,TIFFANY,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ASHCOM,TIFFANY
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 22:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:06

SUBJECT'S LAST NAME: AMARAL

DATE/TIME OF TEST: 01/08/2016 23:32:43

Diagnostics Check		OK	23:26
Air Blank		0.000	23:27
Control Test		0.080	23:27
Air Blank		0.000	23:27
Subject Sample	# 1	0.100	23:28
Breath Volume		1.781	23:28
Air Blank		0.000	23:28
Air Blank		0.000	23:30
Subject Sample	# 2	0.106	23:31
Breath Volume		1.265	23:31
Air Blank		0.000	23:31
Control Test		0.080	23:32
Air Blank		0.000	23:32
Diagnostics Check		OK	23:32

Data Download Date/Time: 02/01/2016 07:03

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BERGER,C
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:39

SUBJECT'S LAST NAME: NIKOLOFF

DATE/TIME OF TEST: 01/09/2016 01:05:44

Diagnostics Check		OK	00:59
Air Blank		0.000	00:59
Control Test		0.080	00:59
Air Blank		0.000	01:00
Subject Sample	# 1	0.205	01:01
Breath Volume		2.023	01:01
Air Blank		0.000	01:01
Air Blank		0.000	01:03
Subject Sample	# 2	0.199	01:04
Breath Volume		1.593	01:04
Air Blank		0.000	01:04
Control Test		0.079	01:05
Air Blank		0.000	01:05
Diagnostics Check		OK	01:05

Data Download Date/Time: 02/01/2016 07:04

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: FELICETTA,F
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 00:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:14

SUBJECT'S LAST NAME: PERROTTE

DATE/TIME OF TEST: 01/09/2016 01:37:43

Diagnostics Check		OK	01:34	
Air Blank		0.000	01:35	
Control Test		0.081	01:35	
Air Blank		0.000	01:36	
Subject Sample	# 1		01:36	Subject Test Refused
Breath Volume		0.000	01:36	
Air Blank		0.000	01:36	
Control Test		0.080	01:37	
Air Blank		0.000	01:37	
Diagnostics Check		OK	01:37	

Data Download Date/Time: 02/01/2016 07:04

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HARR,J
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 01:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: D AREZZO

DATE/TIME OF TEST: 01/09/2016 01:57:38

Diagnostics Check		OK	01:52	
Air Blank		0.000	01:53	
Control Test		0.081	01:53	
Air Blank		0.000	01:53	
Subject Sample	# 1		01:56	Subject Test Refused
Breath Volume		0.000	01:56	
Air Blank		0.000	01:56	
Control Test		0.080	01:57	
Air Blank		0.000	01:57	
Diagnostics Check		OK	01:57	

Data Download Date/Time: 02/01/2016 07:04

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MCDOWELL,D
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:53

SUBJECT'S LAST NAME: FREEMAN

DATE/TIME OF TEST: 01/09/2016 03:16:02

Diagnostics Check		OK	03:13	
Air Blank		0.000	03:13	
Control Test		0.080	03:13	
Air Blank		0.000	03:14	
Subject Sample	# 1		03:14	Subject Test Refused
Breath Volume		0.000	03:14	
Air Blank		0.000	03:15	
Control Test		0.079	03:15	
Air Blank		0.000	03:15	
Diagnostics Check		OK	03:16	

Data Download Date/Time: 02/01/2016 07:04

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEXANDER,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ELLIS,ALEXANDER
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:47

SUBJECT'S LAST NAME: HENSON

DATE/TIME OF TEST: 01/09/2016 04:16:34

Diagnostics Check		OK	04:13	
Air Blank		0.000	04:14	
Control Test		0.081	04:14	
Air Blank		0.000	04:14	
Subject Sample	# 1		04:15	Subject Test Refused
Breath Volume		0.000	04:15	
Air Blank		0.000	04:15	
Control Test		0.081	04:15	
Air Blank		0.000	04:16	
Diagnostics Check		OK	04:16	

Data Download Date/Time: 02/01/2016 07:04

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WEISTOCK,A
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 20:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:45

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 01/09/2016 21:12:20

Diagnostics Check		OK	21:05
Air Blank		0.000	21:05
Control Test		0.081	21:06
Air Blank		0.000	21:06
Subject Sample	# 1	0.155	21:07
Breath Volume		2.121	21:07
Air Blank		0.000	21:08
Air Blank		0.000	21:10
Subject Sample	# 2	0.167	21:10
Breath Volume		1.871	21:10
Air Blank		0.000	21:11
Control Test		0.080	21:11
Air Blank		0.000	21:12
Diagnostics Check		OK	21:12

Data Download Date/Time: 02/01/2016 07:04

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HOOD,C
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 22:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:58

SUBJECT'S LAST NAME: AKIN

DATE/TIME OF TEST: 01/09/2016 23:27:44

Diagnostics Check		OK	23:19
Air Blank		0.000	23:20
Control Test		0.082	23:20
Air Blank		0.000	23:20
Subject Sample	# 1	0.108	23:22
Breath Volume		1.675	23:22
Air Blank		0.000	23:23
Air Blank		0.000	23:25
Subject Sample	# 2	0.108	23:26
Breath Volume		1.488	23:26
Air Blank		0.000	23:26
Control Test		0.080	23:27
Air Blank		0.000	23:27
Diagnostics Check		OK	23:27

Data Download Date/Time: 02/01/2016 07:04

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ASHCOM,TIFFANY,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ASHCOM,TIFFANY
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 02:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:11

SUBJECT'S LAST NAME: MCTIER

DATE/TIME OF TEST: 01/10/2016 03:38:08

Diagnostics Check		OK	03:31
Air Blank		0.000	03:32
Control Test		0.081	03:32
Air Blank		0.000	03:33
Subject Sample	# 1	0.116	03:33
Breath Volume		1.765	03:33
Air Blank		0.000	03:34
Air Blank		0.000	03:36
Subject Sample	# 2	0.117	03:36
Breath Volume		1.531	03:36
Air Blank		0.000	03:37
Control Test		0.080	03:37
Air Blank		0.000	03:38
Diagnostics Check		OK	03:38

Data Download Date/Time: 02/01/2016 07:04

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WEDE,B
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 03:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: BEATON

DATE/TIME OF TEST: 01/10/2016 04:15:05

Diagnostics Check		OK	04:08
Air Blank		0.000	04:09
Control Test		0.081	04:09
Air Blank		0.000	04:10
Subject Sample	# 1	0.169	04:10
Breath Volume		2.378	04:10
Air Blank		0.000	04:11
Air Blank		0.000	04:13
Subject Sample	# 2	0.171	04:13
Breath Volume		2.199	04:13
Air Blank		0.000	04:14
Control Test		0.080	04:14
Air Blank		0.000	04:14
Diagnostics Check		OK	04:15

Data Download Date/Time: 02/01/2016 07:04

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEXANDER,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ELLIS,ALEXANDER
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 01:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: HOLTS

DATE/TIME OF TEST: 01/12/2016 03:03:27

Diagnostics Check		OK	02:57
Air Blank		0.000	02:57
Control Test		0.081	02:58
Air Blank		0.000	02:58
Subject Sample	# 1	0.195	02:59
Breath Volume		1.539	02:59
Air Blank		0.000	02:59
Air Blank		0.000	03:01
Subject Sample	# 2	0.190	03:01
Breath Volume		1.476	03:01
Air Blank		0.000	03:02
Control Test		0.082	03:02
Air Blank		0.000	03:03
Diagnostics Check		OK	03:03

Data Download Date/Time: 02/01/2016 07:04

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ASHCOM,TIFFANY,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ASHCOM,TIFFANY
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 23:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:09

SUBJECT'S LAST NAME: BENDER

DATE/TIME OF TEST: 01/14/2016 00:39:11

Diagnostics Check		OK	00:36	
Air Blank		0.000	00:36	
Control Test		0.082	00:37	
Air Blank		0.000	00:37	
Subject Sample	# 1		00:37	Subject Test Refused
Breath Volume		0.000	00:37	
Air Blank		0.000	00:38	
Control Test		0.083	00:38	
Air Blank		0.000	00:39	
Diagnostics Check		OK	00:39	

Data Download Date/Time: 02/01/2016 07:04

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAGNER,WAYNE,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WAGNER,WAYNE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 23:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:46

SUBJECT'S LAST NAME: CAPITOSTI

DATE/TIME OF TEST: 01/15/2016 00:09:15

Diagnostics Check		OK	00:06	
Air Blank		0.000	00:06	
Control Test		0.081	00:07	
Air Blank		0.000	00:07	
Subject Sample	# 1		00:07	Subject Test Refused
Breath Volume		0.000	00:07	
Air Blank		0.000	00:08	
Control Test		0.082	00:08	
Air Blank		0.000	00:09	
Diagnostics Check		OK	00:09	

Data Download Date/Time: 02/01/2016 07:04

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ASHCOM,TIFFANY,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ASHCOM,TIFFANY
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 21:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:13

SUBJECT'S LAST NAME: MANTER

DATE/TIME OF TEST: 01/15/2016 22:42:30

Diagnostics Check		OK	22:36
Air Blank		0.000	22:36
Control Test		0.081	22:37
Air Blank		0.000	22:37
Subject Sample	# 1	0.274	22:38
Breath Volume		1.226	22:38
Air Blank		0.000	22:38
Air Blank		0.000	22:40
Subject Sample	# 2	0.272	22:40
Breath Volume		1.539	22:40
Air Blank		0.000	22:41
Control Test		0.081	22:41
Air Blank		0.000	22:42
Diagnostics Check		OK	22:42

Data Download Date/Time: 02/01/2016 07:04

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ASHCOM,TIFFANY,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ASHCOM,TIFFANY
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 01:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:07

SUBJECT'S LAST NAME: FORONDA

DATE/TIME OF TEST: 01/16/2016 02:32:06

Diagnostics Check		OK	02:29	
Air Blank		0.000	02:29	
Control Test		0.080	02:30	
Air Blank		0.000	02:30	
Subject Sample	# 1		02:30	Subject Test Refused
Breath Volume		0.000	02:30	
Air Blank		0.000	02:31	
Control Test		0.080	02:31	
Air Blank		0.000	02:32	
Diagnostics Check		OK	02:32	

Data Download Date/Time: 02/01/2016 07:04

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KASELAK,STEPHEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 03:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:19

SUBJECT'S LAST NAME: BROOKS

DATE/TIME OF TEST: 01/16/2016 04:43:20

Diagnostics Check		OK	04:40	
Air Blank		0.000	04:40	
Control Test		0.080	04:41	
Air Blank		0.000	04:41	
Subject Sample	# 1		04:41	Subject Test Refused
Breath Volume		0.000	04:41	
Air Blank		0.000	04:42	
Control Test		0.080	04:42	
Air Blank		0.000	04:43	
Diagnostics Check		OK	04:43	

Data Download Date/Time: 02/01/2016 07:04

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: SCHMITTENDORF,J
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 02:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: RIVERA SANTIAGO

DATE/TIME OF TEST: 01/17/2016 03:19:14

Diagnostics Check		OK	03:10
Air Blank		0.000	03:10
Control Test		0.080	03:10
Air Blank		0.000	03:11
Subject Sample	# 1	0.160	03:13
Breath Volume		1.632	03:13
Air Blank		0.000	03:13
Air Blank		0.000	03:15
Subject Sample	# 2	0.159	03:17
Breath Volume		1.535	03:17
Air Blank		0.000	03:18
Control Test		0.080	03:18
Air Blank		0.000	03:19
Diagnostics Check		OK	03:19

Data Download Date/Time: 02/01/2016 07:04

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SWANNER,SARAH,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: JULIET,KAYLA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 21:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:49

SUBJECT'S LAST NAME: RUSH

DATE/TIME OF TEST: 01/17/2016 22:16:45

Diagnostics Check		OK	22:10
Air Blank		0.000	22:10
Control Test		0.081	22:10
Air Blank		0.000	22:11
Subject Sample	# 1	0.189	22:12
Breath Volume		2.937	22:12
Air Blank		0.000	22:12
Air Blank		0.000	22:14
Subject Sample	# 2	0.195	22:15
Breath Volume		1.785	22:15
Air Blank		0.000	22:15
Control Test		0.080	22:16
Air Blank		0.000	22:16
Diagnostics Check		OK	22:16

Data Download Date/Time: 02/01/2016 07:04

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PERSAUD,COMPTON,D
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: PERSAUD,COMPTON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 22:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:54

SUBJECT'S LAST NAME: MCELWANEY

DATE/TIME OF TEST: 01/20/2016 23:22:02

Diagnostics Check		OK	23:15
Air Blank		0.000	23:16
Control Test		0.079	23:16
Air Blank		0.000	23:16
Subject Sample	# 1	0.106	23:17
Breath Volume		1.562	23:17
Air Blank		0.000	23:18
Air Blank		0.000	23:20
Subject Sample	# 2	0.100	23:20
Breath Volume		1.289	23:20
Air Blank		0.000	23:21
Control Test		0.081	23:21
Air Blank		0.000	23:21
Diagnostics Check		OK	23:22

Data Download Date/Time: 02/01/2016 07:04

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ASHCOM,TIFFANY,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ASHCOM,TIFFANY
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 00:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:06

SUBJECT'S LAST NAME: ROSA SANTIAGO

DATE/TIME OF TEST: 01/21/2016 01:32:17

Diagnostics Check		OK	01:26
Air Blank		0.000	01:26
Control Test		0.080	01:27
Air Blank		0.000	01:27
Subject Sample	# 1	0.032	01:27
Breath Volume		1.781	01:27
Air Blank		0.000	01:28
Air Blank		0.000	01:30
Subject Sample	# 2	0.036	01:30
Breath Volume		1.664	01:30
Air Blank		0.000	01:31
Control Test		0.080	01:31
Air Blank		0.000	01:32
Diagnostics Check		OK	01:32

Data Download Date/Time: 02/01/2016 07:04

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POULIN,DANIELLE,M
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: POULIN,DANIELLE
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 04:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: WALKER

DATE/TIME OF TEST: 01/21/2016 04:57:14

Diagnostics Check		OK	04:50	
Air Blank		0.000	04:50	
Control Test		0.080	04:51	
Air Blank		0.000	04:51	
Subject Sample	# 1	0.145	04:53	
Breath Volume		1.949	04:53	
Air Blank		0.000	04:53	
Air Blank		0.000	04:55	
Subject Sample	# 2	0.133	04:56	
Breath Volume		1.171	04:56	
Air Blank			04:56	RFI Detect
Air Blank		0.000	04:57	

Data Download Date/Time: 02/01/2016 07:04

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LANEY,DAMON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 18:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:00

SUBJECT'S LAST NAME: GARRETT

DATE/TIME OF TEST: 01/22/2016 19:27:23

Diagnostics Check		OK	19:21
Air Blank		0.000	19:21
Control Test		0.080	19:22
Air Blank		0.000	19:22
Subject Sample	# 1	0.000	19:23
Breath Volume		1.707	19:23
Air Blank		0.000	19:23
Air Blank		0.000	19:25
Subject Sample	# 2	0.000	19:26
Breath Volume		1.515	19:26
Air Blank		0.000	19:26
Control Test		0.081	19:26
Air Blank		0.000	19:27
Diagnostics Check		OK	19:27

Data Download Date/Time: 02/01/2016 07:04

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOHR,K,D
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MOHR,K
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 19:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:28

SUBJECT'S LAST NAME: WINTER

DATE/TIME OF TEST: 01/22/2016 19:56:24

Diagnostics Check		OK	19:48
Air Blank		0.000	19:48
Control Test		0.081	19:49
Air Blank		0.000	19:49
Subject Sample	# 1	0.077	19:51
Breath Volume		2.035	19:51
Air Blank		0.000	19:52
Air Blank		0.000	19:54
Subject Sample	# 2	0.078	19:54
Breath Volume		1.832	19:54
Air Blank		0.000	19:55
Control Test		0.080	19:55
Air Blank		0.000	19:56
Diagnostics Check		OK	19:56

Data Download Date/Time: 02/01/2016 07:04

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PERSAUD,COMPTON,D
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: PERSAUD,COMPTON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 02:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:52

SUBJECT'S LAST NAME: QUINN

DATE/TIME OF TEST: 01/23/2016 03:21:25

Diagnostics Check		OK	03:13
Air Blank		0.000	03:14
Control Test		0.080	03:14
Air Blank		0.000	03:15
Subject Sample	# 1	0.150	03:16
Breath Volume		1.472	03:16
Air Blank		0.000	03:17
Air Blank		0.000	03:19
Subject Sample	# 2	0.157	03:19
Breath Volume		1.316	03:19
Air Blank		0.000	03:20
Control Test		0.080	03:20
Air Blank		0.000	03:21
Diagnostics Check		OK	03:21

Data Download Date/Time: 02/01/2016 07:04

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DORR,RANDALL,S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: DORR,RANDALL
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 06:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:55

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 01/23/2016 07:24:31

Diagnostics Check		OK	07:17
Air Blank		0.000	07:18
Control Test		0.080	07:18
Air Blank		0.000	07:19
Subject Sample	# 1	0.176	07:19
Breath Volume		1.480	07:19
Air Blank		0.000	07:20
Air Blank		0.000	07:22
Subject Sample	# 2	0.188	07:23
Breath Volume		1.425	07:23
Air Blank		0.000	07:23
Control Test		0.080	07:23
Air Blank		0.000	07:24
Diagnostics Check		OK	07:24

Data Download Date/Time: 02/01/2016 07:04

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,NICOLE,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: FALLAHEE,TERRANCE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 10:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:29

SUBJECT'S LAST NAME: SEGAL

DATE/TIME OF TEST: 01/23/2016 11:56:04

Diagnostics Check		OK	11:49
Air Blank		0.000	11:50
Control Test		0.080	11:50
Air Blank		0.000	11:51
Subject Sample	# 1	0.000	11:51
Breath Volume		2.386	11:51
Air Blank		0.000	11:52
Air Blank		0.000	11:54
Subject Sample	# 2	0.000	11:54
Breath Volume		2.398	11:54
Air Blank		0.000	11:55
Control Test		0.081	11:55
Air Blank		0.000	11:55
Diagnostics Check		OK	11:56

Data Download Date/Time: 02/01/2016 07:04

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: VERON,C
ARRESTING OFFICER AGENCY: Treasure Island PD
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 22:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: PANASEVICH

DATE/TIME OF TEST: 01/24/2016 00:07:20

Diagnostics Check		OK	00:00
Air Blank		0.000	00:00
Control Test		0.080	00:01
Air Blank		0.000	00:01
Subject Sample	# 1	0.244	00:02
Breath Volume		2.671	00:02
Air Blank		0.000	00:03
Air Blank		0.000	00:05
Subject Sample	# 2	0.249	00:05
Breath Volume		3.023	00:05
Air Blank		0.000	00:06
Control Test		0.080	00:06
Air Blank		0.000	00:07
Diagnostics Check		OK	00:07

Data Download Date/Time: 02/01/2016 07:04

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SWANNER,SARAH,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MOWATT,NATHAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 20:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: DONNELLY

DATE/TIME OF TEST: 01/24/2016 21:58:14

Diagnostics Check		OK	21:50
Air Blank		0.000	21:51
Control Test		0.081	21:51
Air Blank		0.000	21:52
Subject Sample	# 1	0.000	21:53
Breath Volume		2.054	21:53
Air Blank		0.000	21:54
Air Blank		0.000	21:56
Subject Sample	# 2	0.000	21:56
Breath Volume		2.703	21:56
Air Blank		0.000	21:57
Control Test		0.082	21:57
Air Blank		0.000	21:58
Diagnostics Check		OK	21:58

Data Download Date/Time: 02/01/2016 07:04

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NOVAK,RUSSELL,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: NOVAK,RUSSELL
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 23:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: CORMAN

DATE/TIME OF TEST: 01/26/2016 00:32:54

Diagnostics Check		OK	00:26
Air Blank		0.000	00:27
Control Test		0.080	00:27
Air Blank		0.000	00:27
Subject Sample	# 1	0.133	00:28
Breath Volume		1.976	00:28
Air Blank		0.000	00:29
Air Blank		0.000	00:30
Subject Sample	# 2	0.127	00:31
Breath Volume		2.417	00:31
Air Blank		0.000	00:31
Control Test		0.080	00:32
Air Blank		0.000	00:32
Diagnostics Check		OK	00:32

Data Download Date/Time: 02/01/2016 07:04

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAGNER,WAYNE,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WAGNER,WAYNE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 03:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:18

SUBJECT'S LAST NAME: SHELTON

DATE/TIME OF TEST: 01/28/2016 03:45:30

Diagnostics Check		OK	03:38
Air Blank		0.000	03:38
Control Test		0.079	03:39
Air Blank		0.000	03:39
Subject Sample	# 1	0.141	03:40
Breath Volume		1.328	03:40
Air Blank		0.000	03:41
Air Blank		0.000	03:43
Subject Sample	# 2	0.140	03:43
Breath Volume		1.324	03:43
Air Blank		0.000	03:44
Control Test		0.079	03:44
Air Blank		0.000	03:45
Diagnostics Check		OK	03:45

Data Download Date/Time: 02/01/2016 07:04

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HOOD,C
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 20:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: ADAMS

DATE/TIME OF TEST: 01/29/2016 22:56:03

Diagnostics Check		OK	22:53	
Air Blank		0.000	22:53	
Control Test		0.080	22:54	
Air Blank		0.000	22:54	
Subject Sample	# 1		22:54	Subject Test Refused
Breath Volume		0.000	22:54	
Air Blank		0.000	22:55	
Control Test		0.081	22:55	
Air Blank		0.000	22:55	
Diagnostics Check		OK	22:56	

Data Download Date/Time: 02/01/2016 07:04

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HANNA,J
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 22:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: O\CALLAGHAN

DATE/TIME OF TEST: 01/29/2016 23:57:04

Diagnostics Check		OK	23:50
Air Blank		0.000	23:50
Control Test		0.081	23:51
Air Blank		0.000	23:51
Subject Sample	# 1	0.064	23:52
Breath Volume		2.160	23:52
Air Blank		0.000	23:53
Air Blank		0.000	23:54
Subject Sample	# 2	0.063	23:55
Breath Volume		2.105	23:55
Air Blank		0.000	23:56
Control Test		0.080	23:56
Air Blank		0.000	23:56
Diagnostics Check		OK	23:57

Data Download Date/Time: 02/01/2016 07:04

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REFUSE,RONY,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: REFUSE,RONY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 23:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:39

SUBJECT'S LAST NAME: ECKMAN

DATE/TIME OF TEST: 01/30/2016 01:06:08

Diagnostics Check		OK	00:59
Air Blank		0.000	01:00
Control Test		0.080	01:00
Air Blank		0.000	01:00
Subject Sample	# 1	0.183	01:01
Breath Volume		2.449	01:01
Air Blank		0.000	01:02
Air Blank		0.000	01:04
Subject Sample	# 2	0.184	01:04
Breath Volume		2.796	01:04
Air Blank		0.000	01:05
Control Test		0.082	01:05
Air Blank		0.000	01:06
Diagnostics Check		OK	01:06

Data Download Date/Time: 02/01/2016 07:04

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAGNER,WAYNE,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WAGNER,WAYNE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 22:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:31

SUBJECT'S LAST NAME: PRESCOTT

DATE/TIME OF TEST: 01/30/2016 23:02:00

Diagnostics Check		OK	22:54
Air Blank		0.000	22:55
Control Test		0.080	22:55
Air Blank		0.000	22:55
Subject Sample	# 1	0.174	22:57
Breath Volume		2.117	22:57
Air Blank		0.000	22:57
Air Blank		0.000	22:59
Subject Sample	# 2	0.157	23:00
Breath Volume		1.222	23:00
Air Blank		0.000	23:01
Control Test		0.079	23:01
Air Blank		0.000	23:01
Diagnostics Check		OK	23:02

Data Download Date/Time: 02/01/2016 07:04

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRIFFIN,KYLE,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GRIFFIN,KYLE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 01:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:14

SUBJECT'S LAST NAME: HORST

DATE/TIME OF TEST: 01/31/2016 02:46:57

Diagnostics Check		OK	02:35	
Air Blank		0.000	02:35	
Control Test		0.080	02:36	
Air Blank		0.000	02:36	
Subject Sample	# 1	0.131	02:39	Volume Not Met
Breath Volume		0.953	02:39	
Air Blank		0.000	02:40	
Air Blank		0.000	02:42	
Subject Sample	# 2	0.137	02:42	
Breath Volume		1.882	02:42	
Air Blank		0.000	02:43	
Air Blank		0.000	02:45	
Subject Sample	# 3		02:45	Subject Test Refused
Breath Volume		0.000	02:45	
Air Blank		0.000	02:46	
Control Test		0.078	02:46	
Air Blank		0.000	02:46	
Diagnostics Check		OK	02:46	

Data Download Date/Time: 02/01/2016 07:04

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAGNER,WAYNE,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: SCHMITTENDORF,J
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 04:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:35

SUBJECT'S LAST NAME: BRUMM

DATE/TIME OF TEST: 01/31/2016 05:02:39

Diagnostics Check		OK	04:56
Air Blank		0.000	04:56
Control Test		0.080	04:56
Air Blank		0.000	04:57
Subject Sample	# 1	0.183	04:57
Breath Volume		3.695	04:57
Air Blank		0.000	04:58
Air Blank		0.000	05:00
Subject Sample	# 2	0.192	05:01
Breath Volume		3.609	05:01
Air Blank		0.000	05:01
Control Test		0.081	05:02
Air Blank		0.000	05:02
Diagnostics Check		OK	05:02

Data Download Date/Time: 02/01/2016 07:04

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001004

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANSON,JASON,A
BREATH TEST OPERATOR AGENCY: Titusville PD
ARRESTING OFFICER: HANSON,JASON
ARRESTING OFFICER AGENCY: Titusville PD
LAST AGENCY INSPECTION: 12/01/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG400604
DRY GAS EXPIRY DATE: 01/06/2016
VIOLATION CODE: DUI
ARREST TIME: 19:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:37

SUBJECT'S LAST NAME: JAGOE-ALVES

DATE/TIME OF TEST: 01/04/2016 19:53:44

Diagnostics Check		OK	19:47
Air Blank		0.000	19:47
Control Test		0.081	19:47
Air Blank		0.000	19:48
Subject Sample	# 1	0.141	19:49
Breath Volume		1.996	19:49
Air Blank		0.000	19:49
Air Blank		0.000	19:51
Subject Sample	# 2	0.133	19:52
Breath Volume		1.320	19:52
Air Blank		0.000	19:52
Control Test		0.080	19:53
Air Blank		0.000	19:53
Diagnostics Check		OK	19:53

Data Download Date/Time: 01/22/2016 07:22

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001004

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLIDAY,CHELSEA,M
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: EVANS,ANDREW
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 12/01/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG400604
DRY GAS EXPIRY DATE: 01/06/2016
VIOLATION CODE: DUI
ARREST TIME: 01:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: RICHTER

DATE/TIME OF TEST: 01/06/2016 02:12:01

Diagnostics Check		OK	02:05
Air Blank		0.000	02:06
Control Test		0.081	02:06
Air Blank		0.000	02:07
Subject Sample	# 1	0.183	02:07
Breath Volume		4.203	02:07
Air Blank		0.000	02:08
Air Blank		0.000	02:10
Subject Sample	# 2	0.195	02:10
Breath Volume		3.980	02:10
Air Blank		0.000	02:11
Control Test		0.081	02:11
Air Blank		0.000	02:11
Diagnostics Check		OK	02:12

Data Download Date/Time: 01/22/2016 07:22

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001004

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLIDAY,CHELSEA,M
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: HART,JEREMY
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 12/01/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12014080A1
DRY GAS EXPIRY DATE: 06/01/2016
VIOLATION CODE: DUI
ARREST TIME: 00:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: SALTER

DATE/TIME OF TEST: 01/14/2016 01:11:37

Diagnostics Check		OK	01:05
Air Blank		0.000	01:05
Control Test		0.079	01:05
Air Blank		0.000	01:06
Subject Sample	# 1	0.121	01:07
Breath Volume		1.347	01:07
Air Blank		0.000	01:07
Air Blank		0.000	01:09
Subject Sample	# 2	0.134	01:10
Breath Volume		1.464	01:10
Air Blank		0.000	01:10
Control Test		0.079	01:11
Air Blank		0.000	01:11
Diagnostics Check		OK	01:11

Data Download Date/Time: 01/22/2016 07:22

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001004

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKIN,SEAN,D
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: GLEASON,WILLIAM
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 12/01/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12014080A1
DRY GAS EXPIRY DATE: 06/01/2016
VIOLATION CODE: DUI
ARREST TIME: 21:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:24

SUBJECT'S LAST NAME: GODWIN

DATE/TIME OF TEST: 01/17/2016 22:52:55

Diagnostics Check		OK	22:45
Air Blank		0.000	22:46
Control Test		0.079	22:46
Air Blank		0.000	22:46
Subject Sample	# 1	0.272	22:48
Breath Volume		2.500	22:48
Air Blank		0.000	22:48
Air Blank		0.000	22:50
Subject Sample	# 2	0.264	22:51
Breath Volume		1.996	22:51
Air Blank		0.000	22:52
Control Test		0.078	22:52
Air Blank		0.000	22:52
Diagnostics Check		OK	22:52

Data Download Date/Time: 01/22/2016 07:22

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001004

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAVN,SUSANNE,M
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: RAVN,SUSANNE
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 12/01/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12014080A1
DRY GAS EXPIRY DATE: 06/01/2016
VIOLATION CODE: DUI
ARREST TIME: 02:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: SEXTON

DATE/TIME OF TEST: 01/21/2016 03:39:41

Diagnostics Check		OK	03:31
Air Blank		0.000	03:32
Control Test		0.080	03:32
Air Blank		0.000	03:32
Subject Sample	# 1	0.008	03:34
Breath Volume		1.144	03:34
Air Blank		0.000	03:35
Air Blank		0.000	03:37
Subject Sample	# 2	0.000	03:38
Breath Volume		1.324	03:38
Air Blank		0.000	03:38
Control Test		0.079	03:39
Air Blank		0.000	03:39
Diagnostics Check		OK	03:39

Data Download Date/Time: 01/22/2016 07:22

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001004

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCGRORY,DIARMUID,J
BREATH TEST OPERATOR AGENCY: Titusville PD
ARRESTING OFFICER: MCGRORY,DIARMUID
ARRESTING OFFICER AGENCY: Titusville PD
LAST AGENCY INSPECTION: 01/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12014080A1
DRY GAS EXPIRY DATE: 06/01/2016
VIOLATION CODE: DUI
ARREST TIME: 20:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: PIO

DATE/TIME OF TEST: 01/23/2016 21:50:33

Diagnostics Check		OK	21:43
Air Blank		0.000	21:43
Control Test		0.078	21:44
Air Blank		0.000	21:44
Subject Sample	# 1	0.164	21:46
Breath Volume		2.703	21:46
Air Blank		0.000	21:46
Air Blank		0.000	21:48
Subject Sample	# 2	0.184	21:49
Breath Volume		2.269	21:49
Air Blank		0.000	21:49
Control Test		0.079	21:49
Air Blank		0.000	21:50
Diagnostics Check		OK	21:50

Data Download Date/Time: 02/27/2016 18:07

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001004

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAVN,SUSANNE,M
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: RAVN,SUSANNE
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 01/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12014080A1
DRY GAS EXPIRY DATE: 06/01/2016
VIOLATION CODE: DUI
ARREST TIME: 02:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: VAZQUEZ

DATE/TIME OF TEST: 01/26/2016 04:06:23

Diagnostics Check		OK	03:53	
Air Blank		0.000	03:53	
Control Test		0.079	03:54	
Air Blank		0.000	03:54	
Subject Sample	# 1	0.162	03:57	Volume Not Met
Breath Volume		0.031	03:57	
Air Blank		0.000	03:58	
Air Blank		0.000	04:00	
Subject Sample	# 2	0.159	04:01	
Breath Volume		1.445	04:01	
Air Blank		0.000	04:02	
Air Blank		0.000	04:03	
Subject Sample	# 3	0.147	04:04	
Breath Volume		1.773	04:04	
Air Blank		0.000	04:05	
Control Test		0.076	04:05	
Air Blank		0.000	04:06	
Diagnostics Check		OK	04:06	

Data Download Date/Time: 02/27/2016 18:07

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001004

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKIN,SEAN,D
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: MAULDON,KEATON
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 01/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12014080A1
DRY GAS EXPIRY DATE: 06/01/2016
VIOLATION CODE: DUI
ARREST TIME: 01:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: NAVE

DATE/TIME OF TEST: 01/27/2016 02:19:41

Diagnostics Check		OK	02:12
Air Blank		0.000	02:12
Control Test		0.079	02:13
Air Blank		0.000	02:13
Subject Sample	# 1	0.136	02:15
Breath Volume		2.113	02:15
Air Blank		0.000	02:15
Air Blank		0.000	02:17
Subject Sample	# 2	0.136	02:18
Breath Volume		4.027	02:18
Air Blank		0.000	02:18
Control Test		0.078	02:19
Air Blank		0.000	02:19
Diagnostics Check		OK	02:19

Data Download Date/Time: 02/27/2016 18:07

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001004

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARSONS,C,
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: NEGRON,ALEX
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 01/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12014080A1
DRY GAS EXPIRY DATE: 06/01/2016
VIOLATION CODE: DUI
ARREST TIME: 16:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:49

SUBJECT'S LAST NAME: TRUEMAN

DATE/TIME OF TEST: 01/29/2016 17:51:50

Diagnostics Check		OK	17:43
Air Blank		0.000	17:43
Control Test		0.080	17:43
Air Blank		0.000	17:44
Subject Sample	# 1	0.263	17:46
Breath Volume		1.601	17:46
Air Blank		0.000	17:46
Air Blank		0.000	17:48
Subject Sample	# 2	0.273	17:50
Breath Volume		1.433	17:50
Air Blank		0.000	17:50
Control Test		0.078	17:51
Air Blank		0.000	17:51
Diagnostics Check		OK	17:51

Data Download Date/Time: 02/27/2016 18:07

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PERSAUD,COMPTON,D
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: PERSAUD,COMPTON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/31/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 23:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:32

SUBJECT'S LAST NAME: CASALE

DATE/TIME OF TEST: 01/01/2016 00:02:25

Diagnostics Check		OK	23:52
Air Blank		0.000	23:53
Control Test		0.080	23:53
Air Blank		0.000	23:53
Subject Sample	# 2	0.202	23:54
Breath Volume		1.195	23:54
Air Blank		0.000	23:54
Air Blank		0.000	23:56
Subject Sample	# 3	0.175	23:57
Breath Volume		1.492	23:57
Air Blank		0.000	23:57
Air Blank		0.000	23:59
Subject Sample	# 1	0.192	00:00
Breath Volume		1.902	00:00
Air Blank		0.000	00:01
Control Test		0.080	00:01
Air Blank		0.000	00:02
Diagnostics Check		OK	00:02

Data Download Date/Time: 02/01/2016 07:09

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PERSAUD,COMPTON,D
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: PERSAUD,COMPTON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/31/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:09

SUBJECT'S LAST NAME: LOSCHIAVO

DATE/TIME OF TEST: 01/01/2016 03:36:28

Diagnostics Check		OK	03:29
Air Blank		0.000	03:29
Control Test		0.080	03:30
Air Blank		0.000	03:30
Subject Sample	# 1	0.079	03:32
Breath Volume		1.617	03:32
Air Blank		0.000	03:32
Air Blank		0.000	03:34
Subject Sample	# 2	0.081	03:34
Breath Volume		1.644	03:34
Air Blank		0.000	03:35
Control Test		0.079	03:35
Air Blank		0.000	03:36
Diagnostics Check		OK	03:36

Data Download Date/Time: 02/01/2016 07:09

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAGNER,WAYNE,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WAGNER,WAYNE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/31/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 03:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:22

SUBJECT'S LAST NAME: KUNNEN

DATE/TIME OF TEST: 01/01/2016 04:55:11

Diagnostics Check		OK	04:52	
Air Blank		0.000	04:52	
Control Test		0.081	04:53	
Air Blank		0.000	04:53	
Subject Sample	# 1		04:53	Subject Test Refused
Breath Volume		0.000	04:53	
Air Blank		0.000	04:54	
Control Test		0.080	04:54	
Air Blank		0.000	04:55	
Diagnostics Check		OK	04:55	

Data Download Date/Time: 02/01/2016 07:09

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NASSIRI,JAMES,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: NASSIRI,JAMES
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/31/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 17:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:20

SUBJECT'S LAST NAME: CRUZ

DATE/TIME OF TEST: 01/01/2016 18:07:15

Diagnostics Check		OK	18:01
Air Blank		0.000	18:01
Control Test		0.080	18:01
Air Blank		0.000	18:02
Subject Sample	# 1	0.077	18:02
Breath Volume		2.027	18:02
Air Blank		0.000	18:03
Air Blank		0.000	18:05
Subject Sample	# 2	0.073	18:05
Breath Volume		1.828	18:05
Air Blank		0.000	18:06
Control Test		0.080	18:06
Air Blank		0.000	18:07
Diagnostics Check		OK	18:07

Data Download Date/Time: 02/01/2016 07:09

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KASELAK,STEPHEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/31/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 20:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:13

SUBJECT'S LAST NAME: BRIDEAU

DATE/TIME OF TEST: 01/01/2016 21:41:49

Diagnostics Check		OK	21:33
Air Blank		0.000	21:33
Control Test		0.079	21:34
Air Blank		0.000	21:34
Subject Sample	# 1	0.156	21:37
Breath Volume		1.292	21:37
Air Blank		0.000	21:37
Air Blank		0.000	21:39
Subject Sample	# 2	0.161	21:40
Breath Volume		1.277	21:40
Air Blank		0.000	21:40
Control Test		0.079	21:41
Air Blank		0.000	21:41
Diagnostics Check		OK	21:41

Data Download Date/Time: 02/01/2016 07:09

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KASELAK,STEPHEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/31/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 23:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:54

SUBJECT'S LAST NAME: CARR

DATE/TIME OF TEST: 01/02/2016 00:18:31

Diagnostics Check		OK	00:14	
Air Blank		0.000	00:15	
Control Test		0.080	00:15	
Air Blank		0.000	00:16	
Subject Sample	# 1		00:17	Subject Test Refused
Breath Volume		0.000	00:17	
Air Blank		0.000	00:17	
Control Test		0.080	00:17	
Air Blank		0.000	00:18	
Diagnostics Check		OK	00:18	

Data Download Date/Time: 02/01/2016 07:09

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAGNER,WAYNE,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WAGNER,WAYNE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/31/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 02:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:47

SUBJECT'S LAST NAME: LONG

DATE/TIME OF TEST: 01/02/2016 03:14:26

Diagnostics Check		OK	03:07
Air Blank		0.000	03:07
Control Test		0.081	03:08
Air Blank		0.000	03:08
Subject Sample	# 1	0.199	03:09
Breath Volume		3.765	03:09
Air Blank		0.000	03:10
Air Blank		0.000	03:12
Subject Sample	# 2	0.201	03:12
Breath Volume		3.613	03:12
Air Blank		0.000	03:13
Control Test		0.079	03:13
Air Blank		0.000	03:14
Diagnostics Check		OK	03:14

Data Download Date/Time: 02/01/2016 07:09

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KASELAK,STEPHEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/31/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 19:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:15

SUBJECT'S LAST NAME: CECCHI

DATE/TIME OF TEST: 01/02/2016 20:41:48

Diagnostics Check		OK	20:35
Air Blank		0.000	20:35
Control Test		0.080	20:36
Air Blank		0.000	20:36
Subject Sample	# 1	0.000	20:37
Breath Volume		2.554	20:37
Air Blank		0.000	20:37
Air Blank		0.000	20:39
Subject Sample	# 2	0.000	20:40
Breath Volume		3.058	20:40
Air Blank		0.000	20:40
Control Test		0.080	20:41
Air Blank		0.000	20:41
Diagnostics Check		OK	20:41

Data Download Date/Time: 02/01/2016 07:09

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REFUSE,RONY,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: REFUSE,RONY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/31/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 20:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:57

SUBJECT'S LAST NAME: CHRISTOPHER

DATE/TIME OF TEST: 01/02/2016 21:26:22

Diagnostics Check		OK	21:17
Air Blank		0.000	21:18
Control Test		0.079	21:18
Air Blank		0.000	21:18
Subject Sample	# 1	0.120	21:21
Breath Volume		1.238	21:21
Air Blank		0.000	21:21
Air Blank		0.000	21:23
Subject Sample	# 2	0.118	21:24
Breath Volume		1.429	21:24
Air Blank		0.000	21:25
Control Test		0.079	21:25
Air Blank		0.000	21:26
Diagnostics Check		OK	21:26

Data Download Date/Time: 02/01/2016 07:09

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PERSAUD,COMPTON,D
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: PERSAUD,COMPTON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/31/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 22:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:04

SUBJECT'S LAST NAME: BROEMMELSICK

DATE/TIME OF TEST: 01/02/2016 23:27:24

Diagnostics Check		OK	23:24	
Air Blank		0.000	23:24	
Control Test		0.080	23:25	
Air Blank		0.000	23:25	
Subject Sample	# 1		23:26	Subject Test Refused
Breath Volume		0.000	23:26	
Air Blank		0.000	23:26	
Control Test		0.080	23:26	
Air Blank		0.000	23:27	
Diagnostics Check		OK	23:27	

Data Download Date/Time: 02/01/2016 07:09

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HORNING,PETER,T
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HORNING,PETER
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/31/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 02:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:46

SUBJECT'S LAST NAME: BOWLING

DATE/TIME OF TEST: 01/03/2016 03:17:07

Diagnostics Check		OK	03:10
Air Blank		0.000	03:10
Control Test		0.080	03:11
Air Blank		0.000	03:11
Subject Sample	# 1	0.148	03:12
Breath Volume		3.070	03:12
Air Blank		0.000	03:12
Air Blank		0.000	03:14
Subject Sample	# 2	0.154	03:15
Breath Volume		2.796	03:15
Air Blank		0.000	03:16
Control Test		0.080	03:16
Air Blank		0.000	03:17
Diagnostics Check		OK	03:17

Data Download Date/Time: 02/01/2016 07:09

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANGLAIS,SHEILA,B
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: RICOTTILLI,JOSH
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/31/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 01:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:26

SUBJECT'S LAST NAME: HICKS

DATE/TIME OF TEST: 01/04/2016 02:58:53

Diagnostics Check		OK	02:49
Air Blank		0.000	02:50
Control Test		0.080	02:50
Air Blank		0.000	02:51
Subject Sample	# 1	0.063	02:53
Breath Volume		1.718	02:53
Air Blank		0.000	02:54
Air Blank		0.000	02:56
Subject Sample	# 2	0.064	02:57
Breath Volume		1.468	02:57
Air Blank		0.000	02:57
Control Test		0.080	02:58
Air Blank		0.000	02:58
Diagnostics Check		OK	02:58

Data Download Date/Time: 02/01/2016 07:09

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANGLAIS,SHEILA,B
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BOONE,ANTHONY
ARRESTING OFFICER AGENCY: Tarpon Springs PD
LAST AGENCY INSPECTION: 12/31/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 12:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: MALLERY

DATE/TIME OF TEST: 01/05/2016 02:01:32

Diagnostics Check		OK	01:54
Air Blank		0.000	01:55
Control Test		0.081	01:55
Air Blank		0.000	01:55
Subject Sample	# 1	0.159	01:56
Breath Volume		1.753	01:56
Air Blank		0.000	01:57
Air Blank		0.000	01:59
Subject Sample	# 2	0.159	02:00
Breath Volume		1.402	02:00
Air Blank		0.000	02:00
Control Test		0.080	02:00
Air Blank		0.000	02:01
Diagnostics Check		OK	02:01

Data Download Date/Time: 02/01/2016 07:09

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALATI,VINCENZO,F
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GALATI,VINCENZO
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/31/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 20:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:46

SUBJECT'S LAST NAME: CHAMBERLAIN

DATE/TIME OF TEST: 01/06/2016 21:23:54

Diagnostics Check		OK	21:20	
Air Blank		0.000	21:21	
Control Test		0.081	21:21	
Air Blank		0.000	21:22	
Subject Sample	# 1		21:22	Subject Test Refused
Breath Volume		0.000	21:22	
Air Blank		0.000	21:23	
Control Test		0.080	21:23	
Air Blank		0.000	21:23	
Diagnostics Check		OK	21:23	

Data Download Date/Time: 02/01/2016 07:09

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PERSAUD,COMPTON,D
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: PERSAUD,COMPTON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/31/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 00:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:43

SUBJECT'S LAST NAME: LEAMING

DATE/TIME OF TEST: 01/08/2016 01:09:31

Diagnostics Check		OK	01:03
Air Blank		0.000	01:03
Control Test		0.081	01:04
Air Blank		0.000	01:04
Subject Sample	# 1	0.119	01:05
Breath Volume		2.039	01:05
Air Blank		0.000	01:05
Air Blank		0.000	01:07
Subject Sample	# 2	0.124	01:08
Breath Volume		1.652	01:08
Air Blank		0.000	01:08
Control Test		0.080	01:08
Air Blank		0.000	01:09
Diagnostics Check		OK	01:09

Data Download Date/Time: 02/01/2016 07:09

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DORR,RANDALL,S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: DORR,RANDALL
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/31/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 01:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: SANDERS

DATE/TIME OF TEST: 01/08/2016 02:30:41

Diagnostics Check		OK	02:23
Air Blank		0.000	02:23
Control Test		0.081	02:24
Air Blank		0.000	02:24
Subject Sample	# 1	0.186	02:25
Breath Volume		1.558	02:25
Air Blank		0.000	02:26
Air Blank		0.000	02:28
Subject Sample	# 2	0.186	02:29
Breath Volume		1.464	02:29
Air Blank		0.000	02:29
Control Test		0.079	02:30
Air Blank		0.000	02:30
Diagnostics Check		OK	02:30

Data Download Date/Time: 02/01/2016 07:09

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REFUSE,RONY,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: REFUSE,RONY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/31/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 18:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:07

SUBJECT'S LAST NAME: PLANCHET

DATE/TIME OF TEST: 01/08/2016 19:33:26

Diagnostics Check		OK	19:27
Air Blank		0.000	19:27
Control Test		0.080	19:28
Air Blank		0.000	19:28
Subject Sample	# 1	0.071	19:29
Breath Volume		2.957	19:29
Air Blank		0.000	19:29
Air Blank		0.000	19:31
Subject Sample	# 2	0.071	19:31
Breath Volume		3.632	19:31
Air Blank		0.000	19:32
Control Test		0.080	19:32
Air Blank		0.000	19:33
Diagnostics Check		OK	19:33

Data Download Date/Time: 02/01/2016 07:09

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KASELAK,STEPHEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/31/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 20:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:31

SUBJECT'S LAST NAME: WALTZ

DATE/TIME OF TEST: 01/08/2016 22:09:52

Diagnostics Check		OK	21:55	
Air Blank		0.000	21:55	
Control Test		0.080	21:55	
Air Blank		0.000	21:56	
Subject Sample	# 1	0.146	21:56	
Breath Volume		1.515	21:56	
Air Blank		0.000	21:57	
Air Blank		0.000	21:59	
Subject Sample	# 2	0.138	22:02	Volume Not Met
Breath Volume		0.988	22:02	
Air Blank		0.000	22:03	
Air Blank		0.000	22:05	
Subject Sample	# 3	0.133	22:08	Volume Not Met
Breath Volume		0.296	22:08	
Air Blank		0.000	22:08	
Control Test		0.078	22:09	
Air Blank		0.000	22:09	
Diagnostics Check		OK	22:09	

Data Download Date/Time: 02/01/2016 07:09

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KASELAK,STEPHEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/31/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 20:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:31

SUBJECT'S LAST NAME: WALTZ

DATE/TIME OF TEST: 01/08/2016 22:25:45

Diagnostics Check		OK	22:11	
Air Blank		0.000	22:11	
Control Test		0.079	22:12	
Air Blank		0.000	22:12	
Subject Sample	# 1	0.137	22:15	Volume Not Met
Breath Volume		1.000	22:15	
Air Blank		0.000	22:16	
Air Blank		0.000	22:18	
Subject Sample	# 2	0.145	22:19	
Breath Volume		1.218	22:19	
Air Blank		0.000	22:20	
Air Blank		0.000	22:22	
Subject Sample	# 3	0.143	22:24	
Breath Volume		1.183	22:24	
Air Blank		0.000	22:24	
Control Test		0.079	22:25	
Air Blank		0.000	22:25	
Diagnostics Check		OK	22:25	

Data Download Date/Time: 02/01/2016 07:09

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POLANCO,KEDWIN,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: POLANCO,KEDWIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/31/2015

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 21:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:46

SUBJECT'S LAST NAME: ZAKRZEWSKI

DATE/TIME OF TEST: 01/08/2016 23:18:17

Diagnostics Check		OK	23:06	
Air Blank		0.000	23:07	
Control Test		0.079	23:07	
Air Blank		0.000	23:07	
Subject Sample	# 1	0.110	23:11	Volume Not Met
Breath Volume		0.320	23:11	
Air Blank		0.000	23:11	
Air Blank		0.000	23:13	
Subject Sample	# 2	0.107	23:16	Volume Not Met
Breath Volume		0.035	23:16	
Air Blank		0.000	23:17	
Control Test		0.077	23:17	
Air Blank		0.000	23:18	
Diagnostics Check		OK	23:18	

Data Download Date/Time: 02/01/2016 07:10

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAGNER,WAYNE,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WAGNER,WAYNE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/31/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 22:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:11

SUBJECT'S LAST NAME: SADOWSKI

DATE/TIME OF TEST: 01/08/2016 23:41:45

Diagnostics Check		OK	23:35
Air Blank		0.000	23:35
Control Test		0.078	23:36
Air Blank		0.000	23:36
Subject Sample	# 1	0.147	23:37
Breath Volume		2.171	23:37
Air Blank		0.000	23:37
Air Blank		0.000	23:39
Subject Sample	# 2	0.136	23:40
Breath Volume		1.566	23:40
Air Blank		0.000	23:40
Control Test		0.079	23:41
Air Blank		0.000	23:41
Diagnostics Check		OK	23:41

Data Download Date/Time: 02/01/2016 07:10

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KASELAK,STEPHEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/31/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 23:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:54

SUBJECT'S LAST NAME: SHOKITANO

DATE/TIME OF TEST: 01/09/2016 00:17:56

Diagnostics Check		OK	00:15	
Air Blank		0.000	00:15	
Control Test		0.079	00:15	
Air Blank		0.000	00:16	
Subject Sample	# 1		00:16	Subject Test Refused
Breath Volume		0.000	00:16	
Air Blank		0.000	00:17	
Control Test		0.079	00:17	
Air Blank		0.000	00:17	
Diagnostics Check		OK	00:17	

Data Download Date/Time: 02/01/2016 07:10

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAGNER,WAYNE,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WAGNER,WAYNE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/31/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:02

SUBJECT'S LAST NAME: ROBERTS

DATE/TIME OF TEST: 01/09/2016 02:37:02

Diagnostics Check		OK	02:25	
Air Blank		0.000	02:25	
Control Test		0.080	02:26	
Air Blank		0.000	02:26	
Subject Sample	# 1	0.189	02:29	Volume Not Met
Breath Volume		0.183	02:29	
Air Blank		0.000	02:30	
Air Blank		0.000	02:32	
Subject Sample	# 2	0.219	02:35	Volume Not Met
Breath Volume		0.140	02:35	
Air Blank		0.000	02:36	
Control Test		0.079	02:36	
Air Blank		0.000	02:36	
Diagnostics Check		OK	02:37	

Data Download Date/Time: 02/01/2016 07:10

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAIR,RONALD,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAIR,RONALD
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/31/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 02:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:27

SUBJECT'S LAST NAME: REDFIELD

DATE/TIME OF TEST: 01/09/2016 03:01:37

Diagnostics Check		OK	02:47	
Air Blank		0.000	02:48	
Control Test		0.078	02:48	
Air Blank		0.000	02:48	
Subject Sample	# 1	0.128	02:51	
Breath Volume		1.175	02:51	
Air Blank		0.000	02:51	
Air Blank		0.000	02:53	
Subject Sample	# 2	0.123	02:56	Slope Not Level
Breath Volume		1.304	02:56	
Air Blank		0.000	02:57	
Air Blank		0.000	02:59	
Subject Sample	# 3	0.126	03:00	
Breath Volume		1.292	03:00	
Air Blank		0.000	03:00	
Control Test		0.078	03:01	
Air Blank		0.000	03:01	
Diagnostics Check		OK	03:01	

Data Download Date/Time: 02/01/2016 07:10

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAIR,RONALD,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAIR,RONALD
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/31/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 04:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:13

SUBJECT'S LAST NAME: HARKINS

DATE/TIME OF TEST: 01/09/2016 05:36:25

Diagnostics Check		OK	05:33	
Air Blank		0.000	05:34	
Control Test		0.081	05:34	
Air Blank		0.000	05:34	
Subject Sample	# 1		05:35	Subject Test Refused
Breath Volume		0.000	05:35	
Air Blank		0.000	05:35	
Control Test		0.080	05:35	
Air Blank		0.000	05:36	
Diagnostics Check		OK	05:36	

Data Download Date/Time: 02/01/2016 07:10

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KASELAK,STEPHEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/31/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 23:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: EINHORN

DATE/TIME OF TEST: 01/10/2016 00:36:59

Diagnostics Check		OK	00:30
Air Blank		0.000	00:31
Control Test		0.081	00:31
Air Blank		0.000	00:31
Subject Sample	# 1	0.268	00:32
Breath Volume		1.652	00:32
Air Blank		0.000	00:33
Air Blank		0.000	00:35
Subject Sample	# 2	0.279	00:35
Breath Volume		1.714	00:35
Air Blank		0.000	00:36
Control Test		0.079	00:36
Air Blank		0.000	00:36
Diagnostics Check		OK	00:36

Data Download Date/Time: 02/01/2016 07:10

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAGNER,WAYNE,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WAGNER,WAYNE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/31/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 01:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:52

SUBJECT'S LAST NAME: BENNETT

DATE/TIME OF TEST: 01/10/2016 02:17:08

Diagnostics Check		OK	02:14	
Air Blank		0.000	02:14	
Control Test		0.080	02:15	
Air Blank		0.000	02:15	
Subject Sample	# 1		02:15	Subject Test Refused
Breath Volume		0.000	02:15	
Air Blank		0.000	02:16	
Control Test		0.080	02:16	
Air Blank		0.000	02:17	
Diagnostics Check		OK	02:17	

Data Download Date/Time: 02/01/2016 07:10

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAGNER,WAYNE,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WAGNER,WAYNE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/31/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 03:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:42

SUBJECT'S LAST NAME: BOUARFA

DATE/TIME OF TEST: 01/10/2016 04:09:20

Diagnostics Check		OK	04:02
Air Blank		0.000	04:03
Control Test		0.079	04:03
Air Blank		0.000	04:04
Subject Sample	# 1	0.204	04:04
Breath Volume		4.074	04:04
Air Blank		0.000	04:05
Air Blank		0.000	04:07
Subject Sample	# 2	0.194	04:07
Breath Volume		2.054	04:07
Air Blank		0.000	04:08
Control Test		0.080	04:08
Air Blank		0.000	04:09
Diagnostics Check		OK	04:09

Data Download Date/Time: 02/01/2016 07:10

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALATI,VINCENZO,F
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KASELAK,STEPHEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/31/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 04:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:44

SUBJECT'S LAST NAME: KAR

DATE/TIME OF TEST: 01/10/2016 05:11:12

Diagnostics Check		OK	05:04
Air Blank		0.000	05:05
Control Test		0.080	05:05
Air Blank		0.000	05:06
Subject Sample	# 1	0.198	05:06
Breath Volume		1.269	05:06
Air Blank		0.000	05:07
Air Blank		0.000	05:09
Subject Sample	# 2	0.187	05:09
Breath Volume		1.519	05:09
Air Blank		0.000	05:10
Control Test		0.080	05:10
Air Blank		0.000	05:11
Diagnostics Check		OK	05:11

Data Download Date/Time: 02/01/2016 07:10

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE,JONATHAN,S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: DAVIDSON,MATTHEW
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/31/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 19:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:48

SUBJECT'S LAST NAME: NIHOFF

DATE/TIME OF TEST: 01/10/2016 20:15:48

Diagnostics Check		OK	20:08
Air Blank		0.000	20:09
Control Test		0.081	20:09
Air Blank		0.000	20:10
Subject Sample	# 1	0.210	20:11
Breath Volume		2.917	20:11
Air Blank		0.000	20:11
Air Blank		0.000	20:13
Subject Sample	# 2	0.211	20:14
Breath Volume		2.914	20:14
Air Blank		0.000	20:14
Control Test		0.079	20:15
Air Blank		0.000	20:15
Diagnostics Check		OK	20:15

Data Download Date/Time: 02/01/2016 07:10

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANGLIS,SHEILA,B
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HAASE,ANDREW
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/31/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 00:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:37

SUBJECT'S LAST NAME: FINELLI

DATE/TIME OF TEST: 01/11/2016 01:11:04

Diagnostics Check		OK	00:59
Air Blank		0.000	01:00
Control Test		0.079	01:00
Air Blank		0.000	01:01
Subject Sample	# 1	0.265	01:03
Breath Volume		2.656	01:03
Air Blank		0.000	01:04
Air Blank		0.000	01:06
Subject Sample	# 2	0.258	01:09
Breath Volume		2.019	01:09
Air Blank		0.000	01:10
Control Test		0.080	01:10
Air Blank		0.000	01:10
Diagnostics Check		OK	01:11

Data Download Date/Time: 02/01/2016 07:10

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANGLAIS,SHEILA,B
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LANGLAIS,SHEILA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/31/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 04:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:45

SUBJECT'S LAST NAME: MCKISSOCK

DATE/TIME OF TEST: 01/12/2016 05:15:13

Diagnostics Check		OK	05:12	
Air Blank		0.000	05:12	
Control Test		0.080	05:13	
Air Blank		0.000	05:13	
Subject Sample	# 1		05:13	Subject Test Refused
Breath Volume		0.000	05:13	
Air Blank		0.000	05:14	
Control Test		0.081	05:14	
Air Blank		0.000	05:15	
Diagnostics Check		OK	05:15	

Data Download Date/Time: 02/01/2016 07:10

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POLANCO,KEDWIN,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: POLANCO,KEDWIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/31/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 19:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:22

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 01/13/2016 20:50:25

Diagnostics Check		OK	20:43
Air Blank		0.000	20:44
Control Test		0.081	20:44
Air Blank		0.000	20:44
Subject Sample	# 1	0.105	20:45
Breath Volume		1.652	20:45
Air Blank		0.000	20:46
Air Blank		0.000	20:48
Subject Sample	# 2	0.101	20:48
Breath Volume		1.761	20:48
Air Blank		0.000	20:49
Control Test		0.080	20:49
Air Blank		0.000	20:50
Diagnostics Check		OK	20:50

Data Download Date/Time: 02/01/2016 07:10

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAGNER,WAYNE,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WAGNER,WAYNE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/31/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 00:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:09

SUBJECT'S LAST NAME: SOCIA III

DATE/TIME OF TEST: 01/14/2016 01:32:29

Diagnostics Check		OK	01:29	
Air Blank		0.000	01:29	
Control Test		0.081	01:30	
Air Blank		0.000	01:30	
Subject Sample	# 1		01:31	Subject Test Refused
Breath Volume		0.000	01:31	
Air Blank		0.000	01:31	
Control Test		0.080	01:31	
Air Blank		0.000	01:32	
Diagnostics Check		OK	01:32	

Data Download Date/Time: 02/01/2016 07:10

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KASELAK,STEPHEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/31/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 02:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:06

SUBJECT'S LAST NAME: JOHANSEN

DATE/TIME OF TEST: 01/14/2016 03:35:07

Diagnostics Check		OK	03:27
Air Blank		0.000	03:27
Control Test		0.080	03:27
Air Blank		0.000	03:28
Subject Sample	# 1	0.142	03:30
Breath Volume		3.058	03:30
Air Blank		0.000	03:31
Air Blank		0.000	03:33
Subject Sample	# 2	0.137	03:33
Breath Volume		2.023	03:33
Air Blank		0.000	03:34
Control Test		0.080	03:34
Air Blank		0.000	03:35
Diagnostics Check		OK	03:35

Data Download Date/Time: 02/01/2016 07:10

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KASELAK,STEPHEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/31/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 22:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:55

SUBJECT'S LAST NAME: PRAMIS

DATE/TIME OF TEST: 01/14/2016 23:27:38

Diagnostics Check		OK	23:21
Air Blank		0.000	23:21
Control Test		0.079	23:21
Air Blank		0.000	23:22
Subject Sample	# 1	0.237	23:23
Breath Volume		1.980	23:23
Air Blank		0.000	23:23
Air Blank		0.000	23:25
Subject Sample	# 2	0.248	23:26
Breath Volume		2.097	23:26
Air Blank		0.000	23:26
Control Test		0.080	23:27
Air Blank		0.000	23:27
Diagnostics Check		OK	23:27

Data Download Date/Time: 02/01/2016 07:10

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KASELAK,STEPHEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/31/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 22:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: SIEBER

DATE/TIME OF TEST: 01/15/2016 23:02:56

Diagnostics Check		OK	22:56
Air Blank		0.000	22:56
Control Test		0.079	22:57
Air Blank		0.000	22:57
Subject Sample	# 1	0.185	22:58
Breath Volume		3.695	22:58
Air Blank		0.000	22:58
Air Blank		0.000	23:00
Subject Sample	# 2	0.182	23:01
Breath Volume		3.992	23:01
Air Blank		0.000	23:02
Control Test		0.080	23:02
Air Blank		0.000	23:02
Diagnostics Check		OK	23:02

Data Download Date/Time: 02/01/2016 07:10

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAIR,RONALD,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAIR,RONALD
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/31/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 02:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:49

SUBJECT'S LAST NAME: HYBNER

DATE/TIME OF TEST: 01/16/2016 03:13:51

Diagnostics Check		OK	03:09	
Air Blank		0.000	03:10	
Control Test		0.080	03:10	
Air Blank		0.000	03:11	
Subject Sample	# 1		03:12	Subject Test Refused
Breath Volume		0.000	03:12	
Air Blank		0.000	03:12	
Control Test		0.081	03:13	
Air Blank		0.000	03:13	
Diagnostics Check		OK	03:13	

Data Download Date/Time: 02/01/2016 07:10

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GALATI,VINCENZO
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/31/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 04:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:13

SUBJECT'S LAST NAME: CUNNINGHAM

DATE/TIME OF TEST: 01/16/2016 05:41:53

Diagnostics Check		OK	05:35
Air Blank		0.000	05:35
Control Test		0.080	05:35
Air Blank		0.000	05:36
Subject Sample	# 1	0.110	05:37
Breath Volume		3.246	05:37
Air Blank		0.000	05:37
Air Blank		0.000	05:39
Subject Sample	# 2	0.108	05:40
Breath Volume		3.152	05:40
Air Blank		0.000	05:40
Control Test		0.080	05:41
Air Blank		0.000	05:41
Diagnostics Check		OK	05:41

Data Download Date/Time: 02/01/2016 07:10

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANGLAIS,SHEILA,B
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LANGLAIS,SHEILA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/31/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 01:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:44

SUBJECT'S LAST NAME: PRICE

DATE/TIME OF TEST: 01/18/2016 02:12:07

Diagnostics Check		OK	02:05
Air Blank		0.000	02:06
Control Test		0.081	02:06
Air Blank		0.000	02:06
Subject Sample	# 1	0.143	02:07
Breath Volume		3.058	02:07
Air Blank		0.000	02:08
Air Blank		0.000	02:10
Subject Sample	# 2	0.143	02:10
Breath Volume		2.414	02:10
Air Blank		0.000	02:11
Control Test		0.081	02:11
Air Blank		0.000	02:12
Diagnostics Check		OK	02:12

Data Download Date/Time: 02/01/2016 07:10

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVILA,JERRY,W
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: DAVILA,JERRY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/31/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 18:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:05

SUBJECT'S LAST NAME: HEINZ

DATE/TIME OF TEST: 01/18/2016 19:35:31

Diagnostics Check		OK	19:26
Air Blank		0.000	19:27
Control Test		0.080	19:27
Air Blank		0.000	19:28
Subject Sample	# 1	0.155	19:30
Breath Volume		3.152	19:30
Air Blank		0.000	19:31
Air Blank		0.000	19:33
Subject Sample	# 2	0.151	19:33
Breath Volume		2.820	19:33
Air Blank		0.000	19:34
Control Test		0.079	19:34
Air Blank		0.000	19:35
Diagnostics Check		OK	19:35

Data Download Date/Time: 02/01/2016 07:10

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANGLAIS,SHEILA,B
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LANGLAIS,SHEILA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/31/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 23:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:09

SUBJECT'S LAST NAME: MILLSAPS

DATE/TIME OF TEST: 01/19/2016 00:41:25

Diagnostics Check		OK	00:31
Air Blank		0.000	00:31
Control Test		0.081	00:31
Air Blank		0.000	00:32
Subject Sample	# 1	0.304	00:33
Breath Volume		1.472	00:33
Air Blank		0.000	00:33
Air Blank		0.000	00:35
Subject Sample	# 2	0.335	00:36
Breath Volume		1.945	00:36
Air Blank		0.000	00:36
Air Blank		0.000	00:38
Subject Sample	# 3	0.315	00:39
Breath Volume		1.898	00:39
Air Blank		0.000	00:40
Control Test		0.080	00:40
Air Blank		0.000	00:41
Diagnostics Check		OK	00:41

Data Download Date/Time: 02/01/2016 07:10

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANGLAIS,SHEILA,B
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: SHEHAN,KARL
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/31/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 01:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:54

SUBJECT'S LAST NAME: COVELL

DATE/TIME OF TEST: 01/19/2016 02:25:45

Diagnostics Check		OK	02:18
Air Blank		0.000	02:18
Control Test		0.080	02:19
Air Blank		0.000	02:19
Subject Sample	# 1	0.021	02:20
Breath Volume		1.316	02:20
Air Blank		0.000	02:21
Air Blank		0.000	02:22
Subject Sample	# 2	0.020	02:24
Breath Volume		1.582	02:24
Air Blank		0.000	02:24
Control Test		0.081	02:25
Air Blank		0.000	02:25
Diagnostics Check		OK	02:25

Data Download Date/Time: 02/01/2016 07:10

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANGLAIS,SHEILA,B
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HAASE,ANDREW
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/31/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 02:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: PAPILLON

DATE/TIME OF TEST: 01/19/2016 02:55:18

Diagnostics Check		OK	02:52	
Air Blank		0.000	02:52	
Control Test		0.080	02:53	
Air Blank		0.000	02:53	
Subject Sample	# 1		02:53	Subject Test Refused
Breath Volume		0.000	02:53	
Air Blank		0.000	02:54	
Control Test		0.081	02:54	
Air Blank		0.000	02:55	
Diagnostics Check		OK	02:55	

Data Download Date/Time: 02/01/2016 07:10

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANGLAIS,SHEILA,B
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BERGER,CHARLES
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/31/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 23:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:46

SUBJECT'S LAST NAME: RHODES

DATE/TIME OF TEST: 01/20/2016 00:13:42

Diagnostics Check		OK	00:10	
Air Blank		0.000	00:11	
Control Test		0.081	00:11	
Air Blank		0.000	00:12	
Subject Sample	# 1		00:12	Subject Test Refused
Breath Volume		0.000	00:12	
Air Blank		0.000	00:12	
Control Test		0.081	00:13	
Air Blank		0.000	00:13	
Diagnostics Check		OK	00:13	

Data Download Date/Time: 02/01/2016 07:10

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACOBY,LARRY,K
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LEIB,MELANIE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/31/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 01:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:38

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 01/20/2016 02:07:07

Diagnostics Check		OK	01:58
Air Blank		0.000	01:58
Control Test		0.081	01:59
Air Blank		0.000	01:59
Subject Sample	# 1	0.236	02:02
Breath Volume		2.773	02:02
Air Blank		0.000	02:03
Air Blank		0.000	02:04
Subject Sample	# 2	0.217	02:05
Breath Volume		2.089	02:05
Air Blank		0.000	02:06
Control Test		0.079	02:06
Air Blank		0.000	02:07
Diagnostics Check		OK	02:07

Data Download Date/Time: 02/01/2016 07:10

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAGNER,WAYNE,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WAGNER,WAYNE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/31/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 22:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: MARSHALL

DATE/TIME OF TEST: 01/20/2016 23:07:02

Diagnostics Check		OK	23:00
Air Blank		0.000	23:00
Control Test		0.081	23:01
Air Blank		0.000	23:01
Subject Sample	# 1	0.255	23:02
Breath Volume		2.250	23:02
Air Blank		0.000	23:03
Air Blank		0.000	23:05
Subject Sample	# 2	0.253	23:05
Breath Volume		1.773	23:05
Air Blank		0.000	23:06
Control Test		0.080	23:06
Air Blank		0.000	23:06
Diagnostics Check		OK	23:07

Data Download Date/Time: 02/01/2016 07:10

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KASELAK,STEPHEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/31/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:06

SUBJECT'S LAST NAME: SCANLON

DATE/TIME OF TEST: 01/21/2016 00:39:48

Diagnostics Check		OK	00:28	
Air Blank		0.000	00:28	
Control Test		0.081	00:28	
Air Blank		0.000	00:29	
Subject Sample	# 1		00:32	No Sample Provided
Breath Volume		0.000	00:32	
Air Blank		0.000	00:33	
Air Blank		0.000	00:35	
Subject Sample	# 2	0.127	00:38	Volume Not Met
Breath Volume		0.015	00:38	
Air Blank		0.000	00:38	
Control Test		0.081	00:39	
Air Blank		0.000	00:39	
Diagnostics Check		OK	00:39	

Data Download Date/Time: 02/01/2016 07:10

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PERSAUD,COMPTON,D
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: PERSAUD,COMPTON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/31/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 04:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:26

SUBJECT'S LAST NAME: HACKETT

DATE/TIME OF TEST: 01/21/2016 04:54:03

Diagnostics Check		OK	04:46
Air Blank		0.000	04:47
Control Test		0.080	04:47
Air Blank		0.000	04:48
Subject Sample	# 1	0.160	04:49
Breath Volume		1.476	04:49
Air Blank		0.000	04:50
Air Blank		0.000	04:51
Subject Sample	# 2	0.152	04:52
Breath Volume		1.230	04:52
Air Blank		0.000	04:53
Control Test		0.080	04:53
Air Blank		0.000	04:53
Diagnostics Check		OK	04:54

Data Download Date/Time: 02/01/2016 07:10

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILEWSKY,PATRICK,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HUFF,ZACHARY
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 12/31/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 15:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:05

SUBJECT'S LAST NAME: TOVES

DATE/TIME OF TEST: 01/21/2016 16:32:14

Diagnostics Check		OK	16:25
Air Blank		0.000	16:26
Control Test		0.080	16:26
Air Blank		0.000	16:26
Subject Sample	# 1	0.125	16:27
Breath Volume		1.820	16:27
Air Blank		0.000	16:27
Air Blank		0.000	16:29
Subject Sample	# 2	0.117	16:30
Breath Volume		1.683	16:30
Air Blank		0.000	16:31
Control Test		0.080	16:31
Air Blank		0.000	16:32
Diagnostics Check		OK	16:32

Data Download Date/Time: 02/01/2016 07:10

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAGNER,WAYNE,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WAGNER,WAYNE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/31/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 22:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: CREECH

DATE/TIME OF TEST: 01/21/2016 23:02:41

Diagnostics Check		OK	22:56
Air Blank		0.000	22:56
Control Test		0.081	22:56
Air Blank		0.000	22:57
Subject Sample	# 1	0.098	22:57
Breath Volume		3.769	22:57
Air Blank		0.000	22:58
Air Blank		0.000	23:00
Subject Sample	# 2	0.098	23:01
Breath Volume		3.835	23:01
Air Blank		0.000	23:01
Control Test		0.079	23:02
Air Blank		0.000	23:02
Diagnostics Check		OK	23:02

Data Download Date/Time: 02/01/2016 07:10

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAIR,RONALD,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAIR,RONALD
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/31/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 01:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:26

SUBJECT'S LAST NAME: SHARP

DATE/TIME OF TEST: 01/22/2016 01:49:45

Diagnostics Check		OK	01:46	
Air Blank		0.000	01:47	
Control Test		0.081	01:47	
Air Blank		0.000	01:48	
Subject Sample	# 1		01:48	Subject Test Refused
Breath Volume		0.000	01:48	
Air Blank		0.000	01:48	
Control Test		0.081	01:49	
Air Blank		0.000	01:49	
Diagnostics Check		OK	01:49	

Data Download Date/Time: 02/01/2016 07:10

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PERSAUD,COMPTON,D
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: PERSAUD,COMPTON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/31/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 01:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:01

SUBJECT'S LAST NAME: GARCIA VELASCO

DATE/TIME OF TEST: 01/22/2016 02:30:32

Diagnostics Check		OK	02:22
Air Blank		0.000	02:22
Control Test		0.081	02:22
Air Blank		0.000	02:23
Subject Sample	# 1	0.206	02:25
Breath Volume		1.171	02:25
Air Blank		0.000	02:26
Air Blank		0.000	02:28
Subject Sample	# 2	0.208	02:28
Breath Volume		1.292	02:28
Air Blank		0.000	02:29
Control Test		0.079	02:29
Air Blank		0.000	02:30
Diagnostics Check		OK	02:30

Data Download Date/Time: 02/01/2016 07:10

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAGNER,WAYNE,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WAGNER,WAYNE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/31/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 22:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:05

SUBJECT'S LAST NAME: BEVILACQUA

DATE/TIME OF TEST: 01/22/2016 23:35:39

Diagnostics Check		OK	23:29
Air Blank		0.000	23:29
Control Test		0.081	23:30
Air Blank		0.000	23:30
Subject Sample	# 1	0.000	23:31
Breath Volume		2.632	23:31
Air Blank		0.000	23:31
Air Blank		0.000	23:33
Subject Sample	# 2	0.000	23:34
Breath Volume		2.792	23:34
Air Blank		0.000	23:34
Control Test		0.081	23:35
Air Blank		0.000	23:35
Diagnostics Check		OK	23:35

Data Download Date/Time: 02/01/2016 07:10

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PERSAUD,COMPTON,D
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: PERSAUD,COMPTON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/31/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 01:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:08

SUBJECT'S LAST NAME: FORD

DATE/TIME OF TEST: 01/23/2016 01:33:10

Diagnostics Check		OK	01:29	
Air Blank		0.000	01:29	
Control Test		0.080	01:30	
Air Blank		0.000	01:30	
Subject Sample	# 1		01:31	Subject Test Refused
Breath Volume		0.000	01:31	
Air Blank		0.000	01:32	
Control Test		0.080	01:32	
Air Blank		0.000	01:33	
Diagnostics Check		OK	01:33	

Data Download Date/Time: 02/01/2016 07:10

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KASELAK,STEPHEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/31/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 01:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:44

SUBJECT'S LAST NAME: WALKER

DATE/TIME OF TEST: 01/23/2016 02:11:07

Diagnostics Check		OK	02:08	
Air Blank		0.000	02:08	
Control Test		0.080	02:08	
Air Blank		0.000	02:09	
Subject Sample	# 1		02:09	Subject Test Refused
Breath Volume		0.000	02:09	
Air Blank		0.000	02:10	
Control Test		0.080	02:10	
Air Blank		0.000	02:11	
Diagnostics Check		OK	02:11	

Data Download Date/Time: 02/01/2016 07:10

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAIR,RONALD,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAIR,RONALD
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/31/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 02:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:29

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 01/23/2016 02:56:28

Diagnostics Check		OK	02:49
Air Blank		0.000	02:50
Control Test		0.080	02:50
Air Blank		0.000	02:50
Subject Sample	# 1	0.155	02:51
Breath Volume		1.628	02:51
Air Blank		0.000	02:52
Air Blank		0.000	02:54
Subject Sample	# 2	0.164	02:54
Breath Volume		1.667	02:54
Air Blank		0.000	02:55
Control Test		0.080	02:55
Air Blank		0.000	02:56
Diagnostics Check		OK	02:56

Data Download Date/Time: 02/01/2016 07:10

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAGNER,WAYNE,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WAGNER,WAYNE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/31/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 04:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:36

SUBJECT'S LAST NAME: PRICE

DATE/TIME OF TEST: 01/23/2016 05:09:32

Diagnostics Check		OK	05:03
Air Blank		0.000	05:03
Control Test		0.080	05:03
Air Blank		0.000	05:04
Subject Sample	# 1	0.140	05:05
Breath Volume		2.011	05:05
Air Blank		0.000	05:05
Air Blank		0.000	05:07
Subject Sample	# 2	0.138	05:08
Breath Volume		1.468	05:08
Air Blank		0.000	05:08
Control Test		0.080	05:08
Air Blank		0.000	05:09
Diagnostics Check		OK	05:09

Data Download Date/Time: 02/01/2016 07:10

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAIR,RONALD,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAIR,RONALD
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/31/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 05:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:35

SUBJECT'S LAST NAME: HOOKS

DATE/TIME OF TEST: 01/23/2016 05:58:15

Diagnostics Check		OK	05:55	
Air Blank		0.000	05:55	
Control Test		0.080	05:56	
Air Blank		0.000	05:56	
Subject Sample	# 1		05:56	Subject Test Refused
Breath Volume		0.000	05:56	
Air Blank		0.000	05:57	
Control Test		0.080	05:57	
Air Blank		0.000	05:58	
Diagnostics Check		OK	05:58	

Data Download Date/Time: 02/01/2016 07:10

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PERSAUD,COMPTON,D
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: PERSAUD,COMPTON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/31/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 22:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:53

SUBJECT'S LAST NAME: OCONNELL

DATE/TIME OF TEST: 01/23/2016 23:21:42

Diagnostics Check		OK	23:18	
Air Blank		0.000	23:19	
Control Test		0.080	23:19	
Air Blank		0.000	23:20	
Subject Sample	# 1		23:20	Subject Test Refused
Breath Volume		0.000	23:20	
Air Blank		0.000	23:20	
Control Test		0.080	23:21	
Air Blank		0.000	23:21	
Diagnostics Check		OK	23:21	

Data Download Date/Time: 02/01/2016 07:10

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAIR,RONALD,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAIR,RONALD
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/31/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:36

SUBJECT'S LAST NAME: SICHTERMAN

DATE/TIME OF TEST: 01/24/2016 01:03:10

Diagnostics Check		OK	00:59	
Air Blank		0.000	00:59	
Control Test		0.081	01:00	
Air Blank		0.000	01:00	
Subject Sample	# 1		01:01	Subject Test Refused
Breath Volume		0.000	01:01	
Air Blank		0.000	01:02	
Control Test		0.080	01:02	
Air Blank		0.000	01:03	
Diagnostics Check		OK	01:03	

Data Download Date/Time: 02/01/2016 07:10

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PERSAUD,COMPTON,D
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: PERSAUD,COMPTON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/31/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 01:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: FOSTER

DATE/TIME OF TEST: 01/24/2016 01:58:40

Diagnostics Check		OK	01:55	
Air Blank		0.000	01:56	
Control Test		0.081	01:56	
Air Blank		0.000	01:57	
Subject Sample	# 1		01:57	Subject Test Refused
Breath Volume		0.000	01:57	
Air Blank		0.000	01:57	
Control Test		0.080	01:58	
Air Blank		0.000	01:58	
Diagnostics Check		OK	01:58	

Data Download Date/Time: 02/01/2016 07:10

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAGNER,WAYNE,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WAGNER,WAYNE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/31/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: PEGUES

DATE/TIME OF TEST: 01/24/2016 02:36:59

Diagnostics Check		OK	02:25	
Air Blank		0.000	02:26	
Control Test		0.078	02:26	
Air Blank		0.000	02:26	
Subject Sample	# 1		02:30	No Sample Provided
Breath Volume		0.000	02:30	
Air Blank		0.000	02:30	
Air Blank		0.000	02:32	
Subject Sample	# 2		02:35	No Sample Provided
Breath Volume		0.000	02:35	
Air Blank		0.000	02:36	
Control Test		0.079	02:36	
Air Blank		0.000	02:36	
Diagnostics Check		OK	02:36	

Data Download Date/Time: 02/01/2016 07:10

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAIR,RONALD,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAIR,RONALD
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/31/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:26

SUBJECT'S LAST NAME: MECCA

DATE/TIME OF TEST: 01/24/2016 02:49:47

Diagnostics Check		OK	02:47	
Air Blank		0.000	02:47	
Control Test		0.080	02:47	
Air Blank		0.000	02:48	
Subject Sample	# 1		02:48	Subject Test Refused
Breath Volume		0.000	02:48	
Air Blank		0.000	02:48	
Control Test		0.080	02:49	
Air Blank		0.000	02:49	
Diagnostics Check		OK	02:49	

Data Download Date/Time: 02/01/2016 07:10

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILEWSKY,PATRICK,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HANNA,JANE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/31/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 19:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:26

SUBJECT'S LAST NAME: SIMMONS

DATE/TIME OF TEST: 01/24/2016 20:49:55

Diagnostics Check		OK	20:46	
Air Blank		0.000	20:47	
Control Test		0.081	20:47	
Air Blank		0.000	20:47	
Subject Sample	# 1		20:48	Subject Test Refused
Breath Volume		0.000	20:48	
Air Blank		0.000	20:49	
Control Test		0.081	20:49	
Air Blank		0.000	20:49	
Diagnostics Check		OK	20:49	

Data Download Date/Time: 02/01/2016 07:10

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HORNING,PETER,T
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HORNING,PETER
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/31/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 01:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:44

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 01/28/2016 02:24:17

Diagnostics Check		OK	02:12	
Air Blank		0.000	02:13	
Control Test		0.078	02:13	
Air Blank		0.000	02:13	
Subject Sample	# 1	0.100	02:17	Volume Not Met
Breath Volume		0.113	02:17	
Air Blank		0.000	02:17	
Air Blank		0.000	02:19	
Subject Sample	# 2	0.079	02:22	Volume Not Met
Breath Volume		0.449	02:22	
Air Blank		0.000	02:23	
Control Test		0.080	02:23	
Air Blank		0.000	02:24	
Diagnostics Check		OK	02:24	

Data Download Date/Time: 02/01/2016 07:10

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HORNING,PETER,T
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HORNING,PETER
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/31/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 01:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:44

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 01/28/2016 02:34:04

Diagnostics Check		OK	02:26
Air Blank		0.000	02:27
Control Test		0.079	02:27
Air Blank		0.000	02:27
Subject Sample	# 1	0.124	02:28
Breath Volume		1.925	02:28
Air Blank		0.000	02:29
Air Blank		0.000	02:31
Subject Sample	# 2	0.112	02:32
Breath Volume		1.886	02:32
Air Blank		0.000	02:33
Control Test		0.080	02:33
Air Blank		0.000	02:33
Diagnostics Check		OK	02:34

Data Download Date/Time: 02/01/2016 07:10

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KASELAK,STEPHEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/31/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 02:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: DEW

DATE/TIME OF TEST: 01/28/2016 03:35:18

Diagnostics Check		OK	03:29
Air Blank		0.000	03:29
Control Test		0.079	03:29
Air Blank		0.000	03:30
Subject Sample	# 1	0.278	03:30
Breath Volume		2.925	03:30
Air Blank		0.000	03:31
Air Blank		0.000	03:33
Subject Sample	# 2	0.269	03:33
Breath Volume		3.128	03:33
Air Blank		0.000	03:34
Control Test		0.078	03:34
Air Blank		0.000	03:35
Diagnostics Check		OK	03:35

Data Download Date/Time: 02/01/2016 07:10

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PERSAUD,COMPTON,D
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: PERSAUD,COMPTON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/31/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 22:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:16

SUBJECT'S LAST NAME: RUNYAN

DATE/TIME OF TEST: 01/28/2016 23:43:09

Diagnostics Check		OK	23:36
Air Blank		0.000	23:37
Control Test		0.081	23:37
Air Blank		0.000	23:38
Subject Sample	# 1	0.144	23:38
Breath Volume		1.996	23:38
Air Blank		0.000	23:39
Air Blank		0.000	23:40
Subject Sample	# 2	0.137	23:41
Breath Volume		1.902	23:41
Air Blank		0.000	23:42
Control Test		0.081	23:42
Air Blank		0.000	23:43
Diagnostics Check		OK	23:43

Data Download Date/Time: 02/01/2016 07:10

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KASELAK,STEPHEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/31/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 23:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:09

SUBJECT'S LAST NAME: BOVA

DATE/TIME OF TEST: 01/29/2016 00:33:16

Diagnostics Check		OK	00:29	
Air Blank		0.000	00:30	
Control Test		0.081	00:30	
Air Blank		0.000	00:31	
Subject Sample	# 1		00:31	Subject Test Refused
Breath Volume		0.000	00:31	
Air Blank		0.000	00:32	
Control Test		0.081	00:32	
Air Blank		0.000	00:33	
Diagnostics Check		OK	00:33	

Data Download Date/Time: 02/01/2016 07:11

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PERSAUD,COMPTON,D
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: PERSAUD,COMPTON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/31/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:43

SUBJECT'S LAST NAME: SORRELL

DATE/TIME OF TEST: 01/29/2016 02:09:12

Diagnostics Check		OK	02:03
Air Blank		0.000	02:03
Control Test		0.081	02:03
Air Blank		0.000	02:04
Subject Sample	# 1	0.157	02:04
Breath Volume		2.281	02:04
Air Blank		0.000	02:05
Air Blank		0.000	02:07
Subject Sample	# 2	0.155	02:07
Breath Volume		2.171	02:07
Air Blank		0.000	02:08
Control Test		0.080	02:08
Air Blank		0.000	02:09
Diagnostics Check		OK	02:09

Data Download Date/Time: 02/01/2016 07:11

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REEVES,MARK,O
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: SAPANARA,M
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/31/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 06:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:27

SUBJECT'S LAST NAME: GULYUK

DATE/TIME OF TEST: 01/29/2016 07:53:07

Diagnostics Check	OK	07:52	
Air Blank		07:52	RFI Detect
Air Blank	0.000	07:53	

Data Download Date/Time: 02/01/2016 07:11

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REEVES,MARK,O
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: SAPANARA,M
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/31/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 06:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:27

SUBJECT'S LAST NAME: GULYUK

DATE/TIME OF TEST: 01/29/2016 07:59:37

Diagnostics Check		OK	07:56	
Air Blank		0.000	07:57	
Control Test		0.081	07:57	
Air Blank		0.000	07:58	
Subject Sample	# 1		07:58	Subject Test Refused
Breath Volume		0.000	07:58	
Air Blank		0.000	07:58	
Control Test		0.081	07:59	
Air Blank		0.000	07:59	
Diagnostics Check		OK	07:59	

Data Download Date/Time: 02/01/2016 07:11

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KASELAK,STEPHEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/31/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 21:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:14

SUBJECT'S LAST NAME: FRISTER

DATE/TIME OF TEST: 01/29/2016 22:42:04

Diagnostics Check		OK	22:35
Air Blank		0.000	22:36
Control Test		0.082	22:36
Air Blank		0.000	22:37
Subject Sample	# 1	0.133	22:37
Breath Volume		1.394	22:37
Air Blank		0.000	22:38
Air Blank		0.000	22:40
Subject Sample	# 2	0.128	22:40
Breath Volume		1.496	22:40
Air Blank		0.000	22:41
Control Test		0.080	22:41
Air Blank		0.000	22:41
Diagnostics Check		OK	22:42

Data Download Date/Time: 02/01/2016 07:11

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HORNING,PETER,T
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HORNING,PETER
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/31/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: MONDRELLA

DATE/TIME OF TEST: 01/30/2016 00:58:29

Diagnostics Check		OK	00:55	
Air Blank		0.000	00:56	
Control Test		0.081	00:56	
Air Blank		0.000	00:56	
Subject Sample	# 1		00:57	Subject Test Refused
Breath Volume		0.000	00:57	
Air Blank		0.000	00:57	
Control Test		0.080	00:57	
Air Blank		0.000	00:58	
Diagnostics Check		OK	00:58	

Data Download Date/Time: 02/01/2016 07:11

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAIR,RONALD,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAIR,RONALD
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/31/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:43

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 01/30/2016 02:21:05

Diagnostics Check		OK	02:13
Air Blank		0.000	02:13
Control Test		0.081	02:13
Air Blank		0.000	02:14
Subject Sample	# 1	0.000	02:15
Breath Volume		1.472	02:15
Air Blank		0.000	02:16
Air Blank		0.000	02:18
Subject Sample	# 2	0.000	02:19
Breath Volume		1.332	02:19
Air Blank		0.000	02:20
Control Test		0.082	02:20
Air Blank		0.000	02:21
Diagnostics Check		OK	02:21

Data Download Date/Time: 02/01/2016 07:11

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALATI,VINCENZO,F
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GALATI,VINCENZO
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/31/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 02:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: NOLAN

DATE/TIME OF TEST: 01/30/2016 03:41:48

Diagnostics Check		OK	03:35
Air Blank		0.000	03:35
Control Test		0.081	03:35
Air Blank		0.000	03:36
Subject Sample	# 1	0.172	03:37
Breath Volume		1.507	03:37
Air Blank		0.000	03:37
Air Blank		0.000	03:39
Subject Sample	# 2	0.176	03:40
Breath Volume		1.382	03:40
Air Blank		0.000	03:40
Control Test		0.081	03:41
Air Blank		0.000	03:41
Diagnostics Check		OK	03:41

Data Download Date/Time: 02/01/2016 07:11

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KASELAK,STEPHEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/31/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 03:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:57

SUBJECT'S LAST NAME: DORNER

DATE/TIME OF TEST: 01/30/2016 04:25:34

Diagnostics Check		OK	04:19
Air Blank		0.000	04:19
Control Test		0.080	04:19
Air Blank		0.000	04:20
Subject Sample	# 1	0.163	04:20
Breath Volume		1.582	04:20
Air Blank		0.000	04:21
Air Blank		0.000	04:23
Subject Sample	# 2	0.176	04:24
Breath Volume		2.972	04:24
Air Blank		0.000	04:24
Control Test		0.081	04:24
Air Blank		0.000	04:25
Diagnostics Check		OK	04:25

Data Download Date/Time: 02/01/2016 07:11

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GALATI,VINCENZO
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/31/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 04:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:57

SUBJECT'S LAST NAME: MARTINEZ CHAVEZ

DATE/TIME OF TEST: 01/30/2016 05:24:55

Diagnostics Check		OK	05:18
Air Blank		0.000	05:19
Control Test		0.081	05:19
Air Blank		0.000	05:19
Subject Sample	# 1	0.199	05:20
Breath Volume		2.558	05:20
Air Blank		0.000	05:20
Air Blank		0.000	05:22
Subject Sample	# 2	0.203	05:23
Breath Volume		1.371	05:23
Air Blank		0.000	05:24
Control Test		0.080	05:24
Air Blank		0.000	05:24
Diagnostics Check		OK	05:24

Data Download Date/Time: 02/01/2016 07:11

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE,JONATHAN,S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAKE,JONATHAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/31/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 18:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:44

SUBJECT'S LAST NAME: BRIMM

DATE/TIME OF TEST: 01/30/2016 19:16:28

Diagnostics Check		OK	19:07	
Air Blank		0.000	19:08	
Control Test		0.080	19:08	
Air Blank		0.000	19:08	
Subject Sample	# 1		19:12	No Sample Provided
Breath Volume		0.000	19:12	
Air Blank		0.000	19:12	
Air Blank		0.000	19:14	
Subject Sample	# 2		19:15	Subject Test Refused
Breath Volume		0.000	19:15	
Air Blank		0.000	19:15	
Control Test		0.081	19:15	
Air Blank		0.000	19:16	
Diagnostics Check		OK	19:16	

Data Download Date/Time: 02/01/2016 07:11

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAGNER,WAYNE,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WAGNER,WAYNE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/31/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 19:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:23

SUBJECT'S LAST NAME: O\DONNELL

DATE/TIME OF TEST: 01/30/2016 20:50:35

Diagnostics Check		OK	20:44
Air Blank		0.000	20:44
Control Test		0.081	20:44
Air Blank		0.000	20:45
Subject Sample	# 1	0.179	20:45
Breath Volume		2.289	20:45
Air Blank		0.000	20:46
Air Blank		0.000	20:48
Subject Sample	# 2	0.190	20:49
Breath Volume		2.113	20:49
Air Blank		0.000	20:49
Control Test		0.080	20:50
Air Blank		0.000	20:50
Diagnostics Check		OK	20:50

Data Download Date/Time: 02/01/2016 07:11

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALATI,VINCENZO,F
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GALATI,VINCENZO
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/31/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 21:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:19

SUBJECT'S LAST NAME: CONSTANTINO

DATE/TIME OF TEST: 01/30/2016 21:46:16

Diagnostics Check		OK	21:39
Air Blank		0.000	21:40
Control Test		0.078	21:40
Air Blank		0.000	21:41
Subject Sample	# 1	0.128	21:41
Breath Volume		1.503	21:41
Air Blank		0.000	21:42
Air Blank		0.000	21:44
Subject Sample	# 2	0.133	21:44
Breath Volume		1.589	21:44
Air Blank		0.000	21:45
Control Test		0.080	21:45
Air Blank		0.000	21:46
Diagnostics Check		OK	21:46

Data Download Date/Time: 02/01/2016 07:11

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NIXON,AARON,M
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: NIXON,AARON
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/31/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 21:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:52

SUBJECT'S LAST NAME: DEHOYOS

DATE/TIME OF TEST: 01/30/2016 22:24:44

Diagnostics Check		OK	22:18
Air Blank		0.000	22:18
Control Test		0.080	22:18
Air Blank		0.000	22:19
Subject Sample	# 1	0.176	22:20
Breath Volume		1.457	22:20
Air Blank		0.000	22:20
Air Blank		0.000	22:22
Subject Sample	# 2	0.187	22:23
Breath Volume		1.507	22:23
Air Blank		0.000	22:23
Control Test		0.081	22:24
Air Blank		0.000	22:24
Diagnostics Check		OK	22:24

Data Download Date/Time: 02/01/2016 07:11

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUDBRINK,BRIAN,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: SUDBRINK,BRIAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 23:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:22

SUBJECT'S LAST NAME: MATTESON

DATE/TIME OF TEST: 01/31/2016 00:45:12

Diagnostics Check		OK	00:42	
Air Blank		0.000	00:42	
Control Test		0.081	00:43	
Air Blank		0.000	00:43	
Subject Sample	# 1		00:43	Subject Test Refused
Breath Volume		0.000	00:43	
Air Blank		0.000	00:44	
Control Test		0.081	00:44	
Air Blank		0.000	00:45	
Diagnostics Check		OK	00:45	

Data Download Date/Time: 02/01/2016 07:11

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOUIS,WALEX,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: LOUIS,WALEX
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 00:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 01/31/2016 01:14:39

Diagnostics Check		OK	01:06
Air Blank		0.000	01:06
Control Test		0.081	01:06
Air Blank		0.000	01:07
Subject Sample	# 1	0.130	01:08
Breath Volume		2.171	01:08
Air Blank		0.000	01:09
Air Blank		0.000	01:10
Subject Sample	# 2	0.123	01:13
Breath Volume		1.152	01:13
Air Blank		0.000	01:13
Control Test		0.079	01:14
Air Blank		0.000	01:14
Diagnostics Check		OK	01:14

Data Download Date/Time: 02/01/2016 07:11

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KASELAK,STEPHEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 00:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:34

SUBJECT'S LAST NAME: ACOSTA

DATE/TIME OF TEST: 01/31/2016 02:01:56

Diagnostics Check		OK	01:55
Air Blank		0.000	01:55
Control Test		0.081	01:56
Air Blank		0.000	01:56
Subject Sample	# 1	0.117	01:57
Breath Volume		1.738	01:57
Air Blank		0.000	01:57
Air Blank		0.000	01:59
Subject Sample	# 2	0.118	02:00
Breath Volume		2.335	02:00
Air Blank		0.000	02:01
Control Test		0.081	02:01
Air Blank		0.000	02:01
Diagnostics Check		OK	02:01

Data Download Date/Time: 02/01/2016 07:11

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAIR,RONALD,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: EASTTY,MARK
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 01:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:53

SUBJECT'S LAST NAME: CABRAL

DATE/TIME OF TEST: 01/31/2016 02:30:30

Diagnostics Check		OK	02:24
Air Blank		0.000	02:24
Control Test		0.081	02:25
Air Blank		0.000	02:25
Subject Sample	# 1	0.156	02:26
Breath Volume		1.394	02:26
Air Blank		0.000	02:26
Air Blank		0.000	02:28
Subject Sample	# 2	0.150	02:28
Breath Volume		1.417	02:28
Air Blank		0.000	02:29
Control Test		0.079	02:29
Air Blank		0.000	02:30
Diagnostics Check		OK	02:30

Data Download Date/Time: 02/01/2016 07:11

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALATI,VINCENZO,F
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GALATI,VINCENZO
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 02:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:49

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 01/31/2016 03:17:19

Diagnostics Check		OK	03:14	
Air Blank		0.000	03:15	
Control Test		0.081	03:15	
Air Blank		0.000	03:15	
Subject Sample	# 1		03:15	Subject Test Refused
Breath Volume		0.000	03:15	
Air Blank		0.000	03:16	
Control Test		0.081	03:16	
Air Blank		0.000	03:17	
Diagnostics Check		OK	03:17	

Data Download Date/Time: 02/01/2016 07:11

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAGNER,WAYNE,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WAGNER,WAYNE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 03:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:32

SUBJECT'S LAST NAME: JACOBS

DATE/TIME OF TEST: 01/31/2016 04:00:43

Diagnostics Check		OK	03:53
Air Blank		0.000	03:54
Control Test		0.080	03:54
Air Blank		0.000	03:54
Subject Sample	# 1	0.091	03:55
Breath Volume		3.234	03:55
Air Blank		0.000	03:56
Air Blank		0.000	03:58
Subject Sample	# 2	0.093	03:59
Breath Volume		3.410	03:59
Air Blank		0.000	03:59
Control Test		0.080	04:00
Air Blank		0.000	04:00
Diagnostics Check		OK	04:00

Data Download Date/Time: 02/01/2016 07:11

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANGLAIS,SHEILA,B
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LANGLAIS,SHEILA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 03:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:14

SUBJECT'S LAST NAME: PRINDLE

DATE/TIME OF TEST: 01/31/2016 04:51:27

Diagnostics Check		OK	04:39	
Air Blank		0.000	04:40	
Control Test		0.079	04:40	
Air Blank		0.000	04:40	
Subject Sample	# 1	0.237	04:44	Volume Not Met
Breath Volume		0.875	04:44	
Air Blank		0.000	04:44	
Air Blank		0.000	04:46	
Subject Sample	# 2	0.245	04:49	Volume Not Met
Breath Volume		0.199	04:49	
Air Blank		0.000	04:50	
Control Test		0.079	04:50	
Air Blank		0.000	04:51	
Diagnostics Check		OK	04:51	

Data Download Date/Time: 02/01/2016 07:11

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANGLAIS,SHEILA,B
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LANGLAIS,SHEILA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 03:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:14

SUBJECT'S LAST NAME: PRINDLE

DATE/TIME OF TEST: 01/31/2016 05:00:39

Diagnostics Check		OK	04:53	
Air Blank		0.000	04:53	
Control Test		0.078	04:54	
Air Blank		0.000	04:54	
Subject Sample	# 1	0.245	04:56	
Breath Volume		1.425	04:56	
Air Blank		0.000	04:57	
Air Blank		0.000	04:58	
Subject Sample	# 2	0.248	04:59	
Breath Volume		1.402	04:59	
Air Blank			05:00	RFI Detect
Air Blank		0.000	05:00	

Data Download Date/Time: 02/01/2016 07:11

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANGLAIS,SHEILA,B
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LANGLAIS,SHEILA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 03:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:14

SUBJECT'S LAST NAME: PRINDLE

DATE/TIME OF TEST: 01/31/2016 05:06:31

Diagnostics Check		OK	05:03	
Air Blank		0.000	05:03	
Control Test		0.080	05:04	
Air Blank		0.000	05:04	
Subject Sample	# 1		05:05	RFI Detect
Breath Volume		0.000	05:05	
Air Blank		0.000	05:05	
Control Test		0.079	05:05	
Air Blank		0.000	05:06	
Diagnostics Check		OK	05:06	

Data Download Date/Time: 02/01/2016 07:11

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANGLAIS,SHEILA,B
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LANGLAIS,SHEILA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 03:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:14

SUBJECT'S LAST NAME: PRINDLE

DATE/TIME OF TEST: 01/31/2016 05:12:42

Diagnostics Check		OK	05:09	
Air Blank		0.000	05:10	
Control Test		0.080	05:10	
Air Blank		0.000	05:11	
Subject Sample	# 1		05:11	Subject Test Refused
Breath Volume		0.000	05:11	
Air Blank		0.000	05:11	
Control Test		0.079	05:12	
Air Blank		0.000	05:12	
Diagnostics Check		OK	05:12	

Data Download Date/Time: 02/01/2016 07:11

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HART,RANDALL,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: FICOCELLI,MICHAEL
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 17:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:24

SUBJECT'S LAST NAME: DOWDY

DATE/TIME OF TEST: 01/31/2016 18:51:21

Diagnostics Check		OK	18:48	
Air Blank		0.000	18:49	
Control Test		0.080	18:49	
Air Blank		0.000	18:49	
Subject Sample	# 1		18:49	Subject Test Refused
Breath Volume		0.000	18:49	
Air Blank		0.000	18:50	
Control Test		0.080	18:50	
Air Blank		0.000	18:51	
Diagnostics Check		OK	18:51	

Data Download Date/Time: 02/01/2016 07:11

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HART,RANDALL,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: SHEHAN,KARL
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 19:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:14

SUBJECT'S LAST NAME: NELSON

DATE/TIME OF TEST: 01/31/2016 20:37:07

Diagnostics Check		OK	20:34	
Air Blank		0.000	20:34	
Control Test		0.080	20:35	
Air Blank		0.000	20:35	
Subject Sample	# 1		20:35	Subject Test Refused
Breath Volume		0.000	20:35	
Air Blank		0.000	20:36	
Control Test		0.080	20:36	
Air Blank		0.000	20:37	
Diagnostics Check		OK	20:37	

Data Download Date/Time: 02/01/2016 07:11

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001006

INSTRUMENT REGISTERED TO: INDIALANTIC P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CONNOR,MICHAEL,A
BREATH TEST OPERATOR AGENCY: Indialantic PD
ARRESTING OFFICER: TORRES,CLIFF
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 12/02/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 00:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: LUCAS

DATE/TIME OF TEST: 01/01/2016 00:58:03

Diagnostics Check		OK	00:51
Air Blank		0.000	00:52
Control Test		0.078	00:52
Air Blank		0.000	00:53
Subject Sample	# 1	0.197	00:53
Breath Volume		2.871	00:53
Air Blank		0.000	00:54
Air Blank		0.000	00:56
Subject Sample	# 2	0.196	00:56
Breath Volume		2.230	00:56
Air Blank		0.000	00:57
Control Test		0.079	00:57
Air Blank		0.000	00:57
Diagnostics Check		OK	00:58

Data Download Date/Time: 01/01/2016 13:50

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001006

INSTRUMENT REGISTERED TO: INDIALANTIC P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CONNOR,MICHAEL,A
BREATH TEST OPERATOR AGENCY: Indialantic PD
ARRESTING OFFICER: SADLER,JASON
ARRESTING OFFICER AGENCY: Melbourne Beach PD
LAST AGENCY INSPECTION: 12/02/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:12

SUBJECT'S LAST NAME: DIDONATO

DATE/TIME OF TEST: 01/01/2016 02:44:36

Diagnostics Check		OK	02:34
Air Blank		0.000	02:35
Control Test		0.079	02:35
Air Blank		0.000	02:36
Subject Sample	# 1	0.199	02:38
Breath Volume		1.343	02:38
Air Blank		0.000	02:39
Air Blank		0.000	02:41
Subject Sample	# 2	0.199	02:43
Breath Volume		1.152	02:43
Air Blank		0.000	02:43
Control Test		0.077	02:44
Air Blank		0.000	02:44
Diagnostics Check		OK	02:44

Data Download Date/Time: 01/01/2016 13:50

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAGER,KEVIN,
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: JACKSON,C
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 12/31/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12014080A1
DRY GAS EXPIRY DATE: 06/01/2016
VIOLATION CODE: DUI
ARREST TIME: 13:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:14

SUBJECT'S LAST NAME: CAMPBELL

DATE/TIME OF TEST: 01/05/2016 13:43:32

Diagnostics Check		OK	13:39	
Air Blank		0.000	13:39	
Control Test		0.078	13:40	
Air Blank		0.000	13:40	
Subject Sample	# 1		13:42	Subject Test Refused
Breath Volume		0.000	13:42	
Air Blank		0.000	13:42	
Control Test		0.079	13:42	
Air Blank		0.000	13:43	
Diagnostics Check		OK	13:43	

Data Download Date/Time: 01/28/2016 20:17

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,CHAD,B
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: NUNEMAKER,JEFF
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 12/31/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12014080A1
DRY GAS EXPIRY DATE: 06/01/2016
VIOLATION CODE: DUI
ARREST TIME: 02:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 01/09/2016 03:58:47

Diagnostics Check		OK	03:52
Air Blank		0.000	03:53
Control Test		0.077	03:53
Air Blank		0.000	03:53
Subject Sample	# 1	0.174	03:54
Breath Volume		2.933	03:54
Air Blank		0.000	03:54
Air Blank		0.000	03:56
Subject Sample	# 2	0.179	03:57
Breath Volume		3.636	03:57
Air Blank		0.000	03:57
Control Test		0.080	03:58
Air Blank		0.000	03:58
Diagnostics Check		OK	03:58

Data Download Date/Time: 01/28/2016 20:17

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MULDERRIG,SEAN,F
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: MULDERRIG,SEAN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 12/31/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12014080A1
DRY GAS EXPIRY DATE: 06/01/2016
VIOLATION CODE: DUI
ARREST TIME: 00:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:01

SUBJECT'S LAST NAME: RAMOS FLORES

DATE/TIME OF TEST: 01/10/2016 01:28:43

Diagnostics Check		OK	01:21
Air Blank		0.000	01:22
Control Test		0.077	01:22
Air Blank		0.000	01:22
Subject Sample	# 1	0.136	01:23
Breath Volume		2.539	01:23
Air Blank		0.000	01:24
Air Blank		0.000	01:26
Subject Sample	# 2	0.124	01:27
Breath Volume		1.460	01:27
Air Blank		0.000	01:27
Control Test		0.077	01:28
Air Blank		0.000	01:28
Diagnostics Check		OK	01:28

Data Download Date/Time: 01/28/2016 20:17

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MELTZER,CHRISTOPHER,J
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: MELTZER,CHRISTOPHER
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/31/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12014080A1
DRY GAS EXPIRY DATE: 06/01/2016
VIOLATION CODE: DUI
ARREST TIME: 03:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:47

SUBJECT'S LAST NAME: RODRIGUEZ-SANCHEZ

DATE/TIME OF TEST: 01/10/2016 04:16:25

Diagnostics Check		OK	04:09
Air Blank		0.000	04:09
Control Test		0.078	04:10
Air Blank		0.000	04:10
Subject Sample	# 1	0.184	04:11
Breath Volume		1.875	04:11
Air Blank		0.000	04:11
Air Blank		0.000	04:13
Subject Sample	# 2	0.171	04:14
Breath Volume		1.222	04:14
Air Blank		0.000	04:15
Control Test		0.080	04:15
Air Blank		0.000	04:16
Diagnostics Check		OK	04:16

Data Download Date/Time: 01/28/2016 20:17

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MELTZER,CHRISTOPHER,J
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: MELTZER,CHRISTOPHER
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/31/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12014080A1
DRY GAS EXPIRY DATE: 06/01/2016
VIOLATION CODE: DUI
ARREST TIME: 01:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:42

SUBJECT'S LAST NAME: ELLISON

DATE/TIME OF TEST: 01/14/2016 02:14:49

Diagnostics Check		OK	02:05
Air Blank		0.000	02:06
Control Test		0.079	02:06
Air Blank		0.000	02:06
Subject Sample	# 1	0.074	02:08
Breath Volume		1.601	02:08
Air Blank		0.000	02:09
Air Blank		0.000	02:11
Subject Sample	# 2	0.067	02:13
Breath Volume		1.156	02:13
Air Blank		0.000	02:13
Control Test		0.077	02:14
Air Blank		0.000	02:14
Diagnostics Check		OK	02:14

Data Download Date/Time: 01/28/2016 20:17

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001009

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIDGEON,ANDREW,W
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: BADMAN-PATON,LIN
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 12/12/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 09:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:20

SUBJECT'S LAST NAME: GIANLUCA

DATE/TIME OF TEST: 01/18/2016 10:54:48

Diagnostics Check		OK	10:48
Air Blank		0.000	10:49
Control Test		0.079	10:49
Air Blank		0.000	10:49
Subject Sample	# 1	0.137	10:50
Breath Volume		1.664	10:50
Air Blank		0.000	10:50
Air Blank		0.000	10:52
Subject Sample	# 2	0.128	10:53
Breath Volume		2.253	10:53
Air Blank		0.000	10:53
Control Test		0.080	10:54
Air Blank		0.000	10:54
Diagnostics Check		OK	10:54

Data Download Date/Time: 01/24/2016 16:13

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001009

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FREY,ROBERT,P
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: NEGRON,JORGE
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 12/12/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 22:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: CHERWINSKI

DATE/TIME OF TEST: 01/19/2016 23:40:39

Diagnostics Check		OK	23:33
Air Blank		0.000	23:33
Control Test		0.080	23:34
Air Blank		0.000	23:34
Subject Sample	# 1	0.103	23:35
Breath Volume		1.371	23:35
Air Blank		0.000	23:36
Air Blank		0.000	23:38
Subject Sample	# 2	0.099	23:39
Breath Volume		2.003	23:39
Air Blank		0.000	23:39
Control Test		0.079	23:40
Air Blank		0.000	23:40
Diagnostics Check		OK	23:40

Data Download Date/Time: 01/24/2016 16:13

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001009

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOORE,GREG,C
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: ELAFI,MIKE
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 12/12/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 17:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:25

SUBJECT'S LAST NAME: PARKINSON

DATE/TIME OF TEST: 01/21/2016 17:54:51

Diagnostics Check		OK	17:48
Air Blank		0.000	17:49
Control Test		0.080	17:49
Air Blank		0.000	17:49
Subject Sample	# 1	0.164	17:50
Breath Volume		3.031	17:50
Air Blank		0.000	17:51
Air Blank		0.000	17:52
Subject Sample	# 2	0.155	17:53
Breath Volume		1.628	17:53
Air Blank		0.000	17:53
Control Test		0.078	17:54
Air Blank		0.000	17:54
Diagnostics Check		OK	17:54

Data Download Date/Time: 01/24/2016 16:13

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001009

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUTZ,EDWIN,J
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: WEST,CHRISTINA
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 12/12/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: LOCKWOOD

DATE/TIME OF TEST: 01/23/2016 01:39:53

Diagnostics Check		OK	01:33
Air Blank		0.000	01:34
Control Test		0.079	01:34
Air Blank		0.000	01:35
Subject Sample	# 1	0.000	01:35
Breath Volume		2.128	01:35
Air Blank		0.000	01:36
Air Blank		0.000	01:38
Subject Sample	# 2	0.000	01:38
Breath Volume		1.875	01:38
Air Blank		0.000	01:38
Control Test		0.079	01:39
Air Blank		0.000	01:39
Diagnostics Check		OK	01:39

Data Download Date/Time: 01/24/2016 16:13

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001009

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEST,CHRISTINA,A
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: PENNINGTON,WILLIAM
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 01/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 01:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:47

SUBJECT'S LAST NAME: RIVERA-AROCHO

DATE/TIME OF TEST: 01/26/2016 02:20:38

Diagnostics Check		OK	02:14
Air Blank		0.000	02:14
Control Test		0.080	02:15
Air Blank		0.000	02:15
Subject Sample	# 1	0.105	02:16
Breath Volume		1.894	02:16
Air Blank		0.000	02:16
Air Blank		0.000	02:18
Subject Sample	# 2	0.103	02:19
Breath Volume		1.906	02:19
Air Blank		0.000	02:19
Control Test		0.078	02:20
Air Blank		0.000	02:20
Diagnostics Check		OK	02:20

Data Download Date/Time: 02/22/2016 19:57

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001009

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEST,CHRISTINA,A
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: PENNINGTON,WILLIAM
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 01/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 02:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:48

SUBJECT'S LAST NAME: RIVERA NIEVES

DATE/TIME OF TEST: 01/28/2016 03:15:25

Diagnostics Check		OK	03:08
Air Blank		0.000	03:09
Control Test		0.079	03:09
Air Blank		0.000	03:09
Subject Sample	# 1	0.169	03:10
Breath Volume		3.851	03:10
Air Blank		0.000	03:11
Air Blank		0.000	03:13
Subject Sample	# 2	0.159	03:13
Breath Volume		3.261	03:13
Air Blank		0.000	03:14
Control Test		0.077	03:14
Air Blank		0.000	03:15
Diagnostics Check		OK	03:15

Data Download Date/Time: 02/22/2016 19:57

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLISS,F,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DUKIC,S I
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 06:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:48

SUBJECT'S LAST NAME: HUNN JR

DATE/TIME OF TEST: 01/01/2016 08:15:55

Diagnostics Check		OK	08:08
Air Blank		0.000	08:09
Control Test		0.079	08:09
Air Blank		0.000	08:10
Subject Sample	# 1	0.105	08:11
Breath Volume		1.101	08:11
Air Blank		0.000	08:11
Air Blank		0.000	08:13
Subject Sample	# 2	0.099	08:14
Breath Volume		1.460	08:14
Air Blank		0.000	08:14
Control Test		0.079	08:15
Air Blank		0.000	08:15
Diagnostics Check		OK	08:15

Data Download Date/Time: 01/15/2016 09:11

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOSTON,DONALD,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BROWN,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: HAMPTON

DATE/TIME OF TEST: 01/05/2016 23:42:40

Diagnostics Check		OK	23:39	
Air Blank		0.000	23:40	
Control Test		0.081	23:40	
Air Blank		0.000	23:40	
Subject Sample	# 1		23:41	Subject Test Refused
Breath Volume		0.000	23:41	
Air Blank		0.000	23:41	
Control Test		0.081	23:42	
Air Blank		0.000	23:42	
Diagnostics Check		OK	23:42	

Data Download Date/Time: 01/15/2016 09:11

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOSTON,DONALD,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MOSLEY,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:09

SUBJECT'S LAST NAME: CATANESE

DATE/TIME OF TEST: 01/06/2016 02:36:36

Diagnostics Check		OK	02:29
Air Blank		0.000	02:30
Control Test		0.080	02:30
Air Blank		0.000	02:31
Subject Sample	# 1	0.142	02:31
Breath Volume		1.535	02:31
Air Blank		0.000	02:32
Air Blank		0.000	02:34
Subject Sample	# 2	0.147	02:34
Breath Volume		1.726	02:34
Air Blank		0.000	02:35
Control Test		0.079	02:35
Air Blank		0.000	02:36
Diagnostics Check		OK	02:36

Data Download Date/Time: 01/15/2016 09:11

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBER,JOSEPH,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: EANES,J.L.
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:37

SUBJECT'S LAST NAME: SHERRILL

DATE/TIME OF TEST: 01/09/2016 02:00:59

Diagnostics Check		OK	01:57	
Air Blank		0.000	01:58	
Control Test		0.079	01:58	
Air Blank		0.000	01:59	
Subject Sample	# 1		01:59	Subject Test Refused
Breath Volume		0.000	01:59	
Air Blank		0.000	02:00	
Control Test		0.079	02:00	
Air Blank		0.000	02:00	
Diagnostics Check		OK	02:00	

Data Download Date/Time: 01/15/2016 09:11

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBER,JOSEPH,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WALDREP,B.A.
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:53

SUBJECT'S LAST NAME: BACHMANN

DATE/TIME OF TEST: 01/09/2016 03:16:28

Diagnostics Check		OK	03:13	
Air Blank		0.000	03:13	
Control Test		0.079	03:14	
Air Blank		0.000	03:14	
Subject Sample	# 1		03:15	Subject Test Refused
Breath Volume		0.000	03:15	
Air Blank		0.000	03:15	
Control Test		0.079	03:15	
Air Blank		0.000	03:16	
Diagnostics Check		OK	03:16	

Data Download Date/Time: 01/15/2016 09:11

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBER,JOSEPH,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MONTGOMERY,K.R.
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:38

SUBJECT'S LAST NAME: USHER

DATE/TIME OF TEST: 01/09/2016 04:01:42

Diagnostics Check		OK	03:58	
Air Blank		0.000	03:58	
Control Test		0.080	03:59	
Air Blank		0.000	03:59	
Subject Sample	# 1	0.009	04:00	
Breath Volume		1.875	04:00	
Air Blank			04:00	RFI Detect
Air Blank		0.000	04:00	
Control Test		0.079	04:01	
Air Blank		0.000	04:01	
Diagnostics Check		OK	04:01	

Data Download Date/Time: 01/15/2016 09:11

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBER,JOSEPH,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MONTGOMERY,K.R.
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:38

SUBJECT'S LAST NAME: USHER

DATE/TIME OF TEST: 01/09/2016 04:09:31

Diagnostics Check		OK	04:03
Air Blank		0.000	04:03
Control Test		0.081	04:04
Air Blank		0.000	04:04
Subject Sample	# 1	0.008	04:04
Breath Volume		1.507	04:04
Air Blank		0.000	04:05
Air Blank		0.000	04:07
Subject Sample	# 2	0.007	04:08
Breath Volume		1.304	04:08
Air Blank		0.000	04:08
Control Test		0.079	04:08
Air Blank		0.000	04:09
Diagnostics Check		OK	04:09

Data Download Date/Time: 01/15/2016 09:12

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBER,JOSEPH,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HOWARD,J
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 05:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:39

SUBJECT'S LAST NAME: TH

DATE/TIME OF TEST: 01/09/2016 07:07:21

Diagnostics Check		OK	06:59
Air Blank		0.000	06:59
Control Test		0.079	07:00
Air Blank		0.000	07:00
Subject Sample	# 1	0.062	07:02
Breath Volume		2.148	07:02
Air Blank		0.000	07:03
Air Blank		0.000	07:04
Subject Sample	# 2	0.060	07:05
Breath Volume		1.714	07:05
Air Blank		0.000	07:06
Control Test		0.080	07:06
Air Blank		0.000	07:07
Diagnostics Check		OK	07:07

Data Download Date/Time: 01/15/2016 09:12

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JAMES,ALFONSO,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FLORES,D
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 07:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:03

SUBJECT'S LAST NAME: MCADAMS

DATE/TIME OF TEST: 01/09/2016 08:30:21

Diagnostics Check		OK	08:23
Air Blank		0.000	08:24
Control Test		0.080	08:24
Air Blank		0.000	08:25
Subject Sample	# 1	0.120	08:25
Breath Volume		2.726	08:25
Air Blank		0.000	08:26
Air Blank		0.000	08:28
Subject Sample	# 2	0.122	08:28
Breath Volume		3.664	08:28
Air Blank		0.000	08:29
Control Test		0.080	08:29
Air Blank		0.000	08:30
Diagnostics Check		OK	08:30

Data Download Date/Time: 01/15/2016 09:12

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBER,JOSEPH,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FISETTE,J.T.
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:57

SUBJECT'S LAST NAME: LYLES

DATE/TIME OF TEST: 01/13/2016 01:25:36

Diagnostics Check		OK	01:17
Air Blank		0.000	01:18
Control Test		0.080	01:18
Air Blank		0.000	01:18
Subject Sample	# 1	0.204	01:20
Breath Volume		2.351	01:20
Air Blank		0.000	01:21
Air Blank		0.000	01:23
Subject Sample	# 2	0.202	01:23
Breath Volume		2.218	01:23
Air Blank		0.000	01:24
Control Test		0.078	01:24
Air Blank		0.000	01:25
Diagnostics Check		OK	01:25

Data Download Date/Time: 01/15/2016 09:12

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBER,JOSEPH,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: TERRELL,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: SAALMANN

DATE/TIME OF TEST: 01/13/2016 02:15:37

Diagnostics Check		OK	02:07
Air Blank		0.000	02:08
Control Test		0.080	02:08
Air Blank		0.000	02:09
Subject Sample	# 1	0.210	02:10
Breath Volume		1.937	02:10
Air Blank		0.000	02:11
Air Blank		0.000	02:13
Subject Sample	# 2	0.226	02:13
Breath Volume		2.523	02:13
Air Blank		0.000	02:14
Control Test		0.079	02:14
Air Blank		0.000	02:15
Diagnostics Check		OK	02:15

Data Download Date/Time: 01/15/2016 09:12

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBER,JOSEPH,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JONES,C.H.
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:39

SUBJECT'S LAST NAME: FITZGIBBON

DATE/TIME OF TEST: 01/13/2016 04:05:54

Diagnostics Check		OK	03:59
Air Blank		0.000	03:59
Control Test		0.080	03:59
Air Blank		0.000	04:00
Subject Sample	# 1	0.279	04:00
Breath Volume		2.937	04:00
Air Blank		0.000	04:01
Air Blank		0.000	04:03
Subject Sample	# 2	0.278	04:04
Breath Volume		3.054	04:04
Air Blank		0.000	04:04
Control Test		0.079	04:05
Air Blank		0.000	04:05
Diagnostics Check		OK	04:05

Data Download Date/Time: 01/15/2016 09:12

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOSTON,DONALD,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BYRNE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:34

SUBJECT'S LAST NAME: ALLEN

DATE/TIME OF TEST: 01/15/2016 04:08:00

Diagnostics Check		OK	04:04	
Air Blank		0.000	04:05	
Control Test		0.080	04:05	
Air Blank		0.000	04:06	
Subject Sample	# 1		04:06	Subject Test Refused
Breath Volume		0.000	04:06	
Air Blank		0.000	04:07	
Control Test		0.080	04:07	
Air Blank		0.000	04:07	
Diagnostics Check		OK	04:08	

Data Download Date/Time: 01/15/2016 09:12

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON,BRADEN,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: RUDOLPH,H
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 01/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:26

SUBJECT'S LAST NAME: KIRTLEY

DATE/TIME OF TEST: 01/15/2016 23:52:52

Diagnostics Check		OK	23:46
Air Blank		0.000	23:46
Control Test		0.079	23:46
Air Blank		0.000	23:47
Subject Sample	# 1	0.248	23:48
Breath Volume		2.210	23:48
Air Blank		0.000	23:48
Air Blank		0.000	23:50
Subject Sample	# 2	0.240	23:51
Breath Volume		1.691	23:51
Air Blank		0.000	23:51
Control Test		0.077	23:52
Air Blank		0.000	23:52
Diagnostics Check		OK	23:52

Data Download Date/Time: 01/29/2016 08:54

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLEY,SHANNON,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ANTHONY,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:50

SUBJECT'S LAST NAME: SIMMONS

DATE/TIME OF TEST: 01/16/2016 07:29:48

Diagnostics Check		OK	07:23
Air Blank		0.000	07:23
Control Test		0.080	07:23
Air Blank		0.000	07:24
Subject Sample	# 1	0.159	07:25
Breath Volume		2.304	07:25
Air Blank		0.000	07:25
Air Blank		0.000	07:27
Subject Sample	# 2	0.156	07:28
Breath Volume		1.734	07:28
Air Blank		0.000	07:28
Control Test		0.078	07:29
Air Blank		0.000	07:29
Diagnostics Check		OK	07:29

Data Download Date/Time: 01/29/2016 08:55

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOSTON,DONALD,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BYRNE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: DIRANI

DATE/TIME OF TEST: 01/17/2016 23:49:31

Diagnostics Check		OK	23:46	
Air Blank		0.000	23:47	
Control Test		0.081	23:47	
Air Blank		0.000	23:47	
Subject Sample	# 1		23:48	Subject Test Refused
Breath Volume		0.000	23:48	
Air Blank		0.000	23:48	
Control Test		0.079	23:48	
Air Blank		0.000	23:49	
Diagnostics Check		OK	23:49	

Data Download Date/Time: 01/29/2016 08:55

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOSTON,DONALD,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MOSLEY,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:33

SUBJECT'S LAST NAME: FLEMING

DATE/TIME OF TEST: 01/18/2016 01:06:06

Diagnostics Check		OK	00:53	
Air Blank		0.000	00:53	
Control Test		0.080	00:54	
Air Blank		0.000	00:54	
Subject Sample	# 1	0.045	00:57	Volume Not Met
Breath Volume		0.101	00:57	
Air Blank		0.000	00:58	
Air Blank		0.000	01:00	
Subject Sample	# 2	0.141	01:01	
Breath Volume		1.207	01:01	
Air Blank		0.000	01:02	
Air Blank		0.000	01:03	
Subject Sample	# 3	0.145	01:04	
Breath Volume		2.324	01:04	
Air Blank		0.000	01:05	
Control Test		0.079	01:05	
Air Blank		0.000	01:06	
Diagnostics Check		OK	01:06	

Data Download Date/Time: 01/29/2016 08:55

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOSTON,DONALD,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: EANES,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: SARROCA

DATE/TIME OF TEST: 01/19/2016 02:43:27

Diagnostics Check		OK	02:36
Air Blank		0.000	02:36
Control Test		0.081	02:37
Air Blank		0.000	02:37
Subject Sample	# 1	0.096	02:38
Breath Volume		1.320	02:38
Air Blank		0.000	02:39
Air Blank		0.000	02:40
Subject Sample	# 2	0.110	02:41
Breath Volume		2.519	02:41
Air Blank		0.000	02:42
Control Test		0.080	02:42
Air Blank		0.000	02:43
Diagnostics Check		OK	02:43

Data Download Date/Time: 01/29/2016 08:55

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTELL,CHRIS,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: GARCIA,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 18:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:38

SUBJECT'S LAST NAME: RIOS DELGADO

DATE/TIME OF TEST: 01/19/2016 20:02:18

Diagnostics Check		OK	19:58	
Air Blank		0.000	19:58	
Control Test		0.081	19:59	
Air Blank		0.000	19:59	
Subject Sample	# 1	0.102	20:00	
Breath Volume		1.332	20:00	
Air Blank			20:00	RFI Detect
Air Blank		0.000	20:01	
Control Test		0.080	20:01	
Air Blank		0.000	20:02	
Diagnostics Check		OK	20:02	

Data Download Date/Time: 01/29/2016 08:55

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTELL,CHRIS,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: GARCIA,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 18:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:38

SUBJECT'S LAST NAME: RIOS DELGADO

DATE/TIME OF TEST: 01/19/2016 20:11:25

Diagnostics Check		OK	20:04
Air Blank		0.000	20:05
Control Test		0.080	20:05
Air Blank		0.000	20:05
Subject Sample	# 1	0.114	20:06
Breath Volume		1.546	20:06
Air Blank		0.000	20:07
Air Blank		0.000	20:08
Subject Sample	# 2	0.123	20:09
Breath Volume		1.316	20:09
Air Blank		0.000	20:10
Control Test		0.079	20:10
Air Blank		0.000	20:11
Diagnostics Check		OK	20:11

Data Download Date/Time: 01/29/2016 08:55

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAY,DANIEL,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DAVENPORT,R
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 01/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 19:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:14

SUBJECT'S LAST NAME: GARLAND

DATE/TIME OF TEST: 01/20/2016 22:38:36

Diagnostics Check		OK	22:35	
Air Blank		0.000	22:35	
Control Test		0.080	22:36	
Air Blank		0.000	22:36	
Subject Sample	# 1		22:37	Subject Test Refused
Breath Volume		0.000	22:37	
Air Blank		0.000	22:37	
Control Test		0.080	22:37	
Air Blank		0.000	22:38	
Diagnostics Check		OK	22:38	

Data Download Date/Time: 01/29/2016 08:55

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBER,JOSEPH,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: GARCIA,J.R.
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: POMPEY

DATE/TIME OF TEST: 01/21/2016 00:49:48

Diagnostics Check		OK	00:46	
Air Blank		0.000	00:47	
Control Test		0.082	00:47	
Air Blank		0.000	00:47	
Subject Sample	# 1		00:48	Subject Test Refused
Breath Volume		0.000	00:48	
Air Blank		0.000	00:48	
Control Test		0.080	00:49	
Air Blank		0.000	00:49	
Diagnostics Check		OK	00:49	

Data Download Date/Time: 01/29/2016 08:55

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBER,JOSEPH,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BARTILUCCI,D.E.
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:03

SUBJECT'S LAST NAME: CARTER

DATE/TIME OF TEST: 01/21/2016 01:30:48

Diagnostics Check		OK	01:24
Air Blank		0.000	01:24
Control Test		0.080	01:24
Air Blank		0.000	01:25
Subject Sample	# 1	0.056	01:25
Breath Volume		3.328	01:25
Air Blank		0.000	01:26
Air Blank		0.000	01:28
Subject Sample	# 2	0.057	01:29
Breath Volume		3.382	01:29
Air Blank		0.000	01:29
Control Test		0.081	01:30
Air Blank		0.000	01:30
Diagnostics Check		OK	01:30

Data Download Date/Time: 01/29/2016 08:55

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBER,JOSEPH,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BOWEN,J.W.
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 01/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:11

SUBJECT'S LAST NAME: HAYDEN

DATE/TIME OF TEST: 01/21/2016 02:38:52

Diagnostics Check		OK	02:32
Air Blank		0.000	02:32
Control Test		0.080	02:32
Air Blank		0.000	02:33
Subject Sample	# 1	0.128	02:33
Breath Volume		1.335	02:33
Air Blank		0.000	02:34
Air Blank		0.000	02:36
Subject Sample	# 2	0.139	02:37
Breath Volume		2.128	02:37
Air Blank		0.000	02:37
Control Test		0.080	02:38
Air Blank		0.000	02:38
Diagnostics Check		OK	02:38

Data Download Date/Time: 01/29/2016 08:55

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBER,JOSEPH,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MITLYNG,T.C.
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:42

SUBJECT'S LAST NAME: COTNER

DATE/TIME OF TEST: 01/21/2016 04:09:11

Diagnostics Check		OK	04:02
Air Blank		0.000	04:03
Control Test		0.080	04:03
Air Blank		0.000	04:03
Subject Sample	# 1	0.110	04:04
Breath Volume		2.734	04:04
Air Blank		0.000	04:05
Air Blank		0.000	04:06
Subject Sample	# 2	0.117	04:07
Breath Volume		3.500	04:07
Air Blank		0.000	04:08
Control Test		0.080	04:08
Air Blank		0.000	04:09
Diagnostics Check		OK	04:09

Data Download Date/Time: 01/29/2016 08:55

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBER,JOSEPH,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: GARCIA,J.R.
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: NIPPER

DATE/TIME OF TEST: 01/21/2016 23:27:15

Diagnostics Check		OK	23:20
Air Blank		0.000	23:20
Control Test		0.081	23:21
Air Blank		0.000	23:21
Subject Sample	# 1	0.267	23:22
Breath Volume		2.695	23:22
Air Blank		0.000	23:23
Air Blank		0.000	23:24
Subject Sample	# 2	0.264	23:25
Breath Volume		2.628	23:25
Air Blank		0.000	23:26
Control Test		0.078	23:26
Air Blank		0.000	23:27
Diagnostics Check		OK	23:27

Data Download Date/Time: 01/29/2016 08:55

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBER,JOSEPH,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: RAMIREZ,I
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 01/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:48

SUBJECT'S LAST NAME: HOGUE

DATE/TIME OF TEST: 01/22/2016 01:15:08

Diagnostics Check		OK	01:08
Air Blank		0.000	01:08
Control Test		0.079	01:09
Air Blank		0.000	01:09
Subject Sample	# 1	0.143	01:10
Breath Volume		4.183	01:10
Air Blank		0.000	01:11
Air Blank		0.000	01:12
Subject Sample	# 2	0.137	01:13
Breath Volume		4.253	01:13
Air Blank		0.000	01:14
Control Test		0.078	01:14
Air Blank		0.000	01:15
Diagnostics Check		OK	01:15

Data Download Date/Time: 01/29/2016 08:55

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBER,JOSEPH,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BARTILUCCI,D
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:44

SUBJECT'S LAST NAME: HASKEW

DATE/TIME OF TEST: 01/22/2016 02:14:08

Diagnostics Check		OK	02:04
Air Blank		0.000	02:05
Control Test		0.078	02:05
Air Blank		0.000	02:06
Subject Sample	# 1	0.207	02:07
Breath Volume		3.238	02:07
Air Blank		0.000	02:08
Air Blank		0.000	02:09
Subject Sample	# 2	0.201	02:12
Breath Volume		3.785	02:12
Air Blank		0.000	02:13
Control Test		0.075	02:13
Air Blank		0.000	02:14
Diagnostics Check		OK	02:14

Data Download Date/Time: 01/29/2016 08:55

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBER,JOSEPH,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FISETTE,J.T.
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:36

SUBJECT'S LAST NAME: MAZZA

DATE/TIME OF TEST: 01/22/2016 03:00:01

Diagnostics Check		OK	02:56	
Air Blank		0.000	02:57	
Control Test		0.078	02:57	
Air Blank		0.000	02:58	
Subject Sample	# 1		02:58	Subject Test Refused
Breath Volume		0.000	02:58	
Air Blank		0.000	02:59	
Control Test		0.079	02:59	
Air Blank		0.000	02:59	
Diagnostics Check		OK	03:00	

Data Download Date/Time: 01/29/2016 08:55

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBER,JOSEPH,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: GARCIA,J.R.
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:54

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 01/22/2016 23:18:41

Diagnostics Check		OK	23:14	
Air Blank		0.000	23:15	
Control Test		0.080	23:15	
Air Blank		0.000	23:16	
Subject Sample	# 1		23:17	Subject Test Refused
Breath Volume		0.000	23:17	
Air Blank		0.000	23:17	
Control Test		0.078	23:18	
Air Blank		0.000	23:18	
Diagnostics Check		OK	23:18	

Data Download Date/Time: 01/29/2016 08:55

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBER,JOSEPH,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FISETTE,J.T.
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:48

SUBJECT'S LAST NAME: HART

DATE/TIME OF TEST: 01/23/2016 00:17:11

Diagnostics Check		OK	00:10
Air Blank		0.000	00:11
Control Test		0.078	00:11
Air Blank		0.000	00:11
Subject Sample	# 1	0.061	00:12
Breath Volume		4.777	00:12
Air Blank		0.000	00:13
Air Blank		0.000	00:14
Subject Sample	# 2	0.060	00:15
Breath Volume		4.320	00:15
Air Blank		0.000	00:16
Control Test		0.079	00:16
Air Blank		0.000	00:17
Diagnostics Check		OK	00:17

Data Download Date/Time: 01/29/2016 08:55

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBER,JOSEPH,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: GARCIA,J.R.
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:47

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 01/23/2016 02:10:52

Diagnostics Check		OK	02:07	
Air Blank		0.000	02:08	
Control Test		0.080	02:08	
Air Blank		0.000	02:09	
Subject Sample	# 1		02:09	Subject Test Refused
Breath Volume		0.000	02:09	
Air Blank		0.000	02:09	
Control Test		0.079	02:10	
Air Blank		0.000	02:10	
Diagnostics Check		OK	02:10	

Data Download Date/Time: 01/29/2016 08:55

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBER,JOSEPH,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ALBERT,B.T.
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:35

SUBJECT'S LAST NAME: BOUNCE

DATE/TIME OF TEST: 01/23/2016 04:59:12

Diagnostics Check		OK	04:56	
Air Blank		0.000	04:56	
Control Test		0.080	04:56	
Air Blank		0.000	04:57	
Subject Sample	# 1		04:57	Subject Test Refused
Breath Volume		0.000	04:57	
Air Blank		0.000	04:58	
Control Test		0.079	04:58	
Air Blank		0.000	04:59	
Diagnostics Check		OK	04:59	

Data Download Date/Time: 01/29/2016 08:55

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBER,JOSEPH,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: PURVIS,E
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 01/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:00

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 01/23/2016 06:25:41

Diagnostics Check		OK	06:21	
Air Blank		0.000	06:21	
Control Test		0.081	06:22	
Air Blank		0.000	06:22	
Subject Sample	# 1		06:24	Subject Test Refused
Breath Volume		0.000	06:24	
Air Blank		0.000	06:24	
Control Test		0.080	06:25	
Air Blank		0.000	06:25	
Diagnostics Check		OK	06:25	

Data Download Date/Time: 01/29/2016 08:55

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBER,JOSEPH,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WALDREP,B.A.
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 01/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:43

SUBJECT'S LAST NAME: MORGAN

DATE/TIME OF TEST: 01/25/2016 04:11:34

Diagnostics Check		OK	04:04
Air Blank		0.000	04:05
Control Test		0.081	04:05
Air Blank		0.000	04:05
Subject Sample	# 1	0.233	04:06
Breath Volume		3.011	04:06
Air Blank		0.000	04:07
Air Blank		0.000	04:08
Subject Sample	# 2	0.230	04:09
Breath Volume		3.160	04:09
Air Blank		0.000	04:10
Control Test		0.079	04:10
Air Blank		0.000	04:11
Diagnostics Check		OK	04:11

Data Download Date/Time: 01/29/2016 08:55

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBER,JOSEPH,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BAI,S.W.
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:11

SUBJECT'S LAST NAME: RAMOS ULLOA

DATE/TIME OF TEST: 01/26/2016 01:38:24

Diagnostics Check		OK	01:31
Air Blank		0.000	01:31
Control Test		0.080	01:32
Air Blank		0.000	01:32
Subject Sample	# 1	0.211	01:33
Breath Volume		1.257	01:33
Air Blank		0.000	01:34
Air Blank		0.000	01:35
Subject Sample	# 2	0.214	01:36
Breath Volume		2.835	01:36
Air Blank		0.000	01:37
Control Test		0.079	01:37
Air Blank		0.000	01:38
Diagnostics Check		OK	01:38

Data Download Date/Time: 01/29/2016 08:55

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOSTON,DONALD,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MOSLEY,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: COATES

DATE/TIME OF TEST: 01/27/2016 02:19:13

Diagnostics Check		OK	02:16	
Air Blank		0.000	02:16	
Control Test		0.080	02:17	
Air Blank		0.000	02:17	
Subject Sample	# 1		02:17	Subject Test Refused
Breath Volume		0.000	02:17	
Air Blank		0.000	02:18	
Control Test		0.079	02:18	
Air Blank		0.000	02:19	
Diagnostics Check		OK	02:19	

Data Download Date/Time: 01/29/2016 08:55

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOSTON,DONALD,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: COLEMAN,F
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:23

SUBJECT'S LAST NAME: POLKE

DATE/TIME OF TEST: 01/27/2016 23:49:16

Diagnostics Check		OK	23:44	
Air Blank		0.000	23:44	
Control Test		0.080	23:45	
Air Blank		0.000	23:45	
Subject Sample	# 1		23:47	Subject Test Refused
Breath Volume		0.000	23:47	
Air Blank		0.000	23:48	
Control Test		0.079	23:48	
Air Blank		0.000	23:49	
Diagnostics Check		OK	23:49	

Data Download Date/Time: 01/29/2016 08:55

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOSTON,DONALD,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MONTGOMERY,K
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 01/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:38

SUBJECT'S LAST NAME: COCANOUGHER

DATE/TIME OF TEST: 01/28/2016 04:01:29

Diagnostics Check		OK	03:59	
Air Blank		0.000	03:59	
Control Test		0.080	03:59	
Air Blank		0.000	04:00	
Subject Sample	# 1		04:00	Subject Test Refused
Breath Volume		0.000	04:00	
Air Blank			04:01	RFI Detect
Air Blank		0.000	04:01	

Data Download Date/Time: 01/29/2016 08:55

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOSTON,DONALD,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MONTGOMERY,K
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 01/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:38

SUBJECT'S LAST NAME: COCANOUGHER

DATE/TIME OF TEST: 01/28/2016 04:06:04

Diagnostics Check		OK	04:03	
Air Blank		0.000	04:03	
Control Test		0.080	04:03	
Air Blank		0.000	04:04	
Subject Sample	# 1		04:04	Subject Test Refused
Breath Volume		0.000	04:04	
Air Blank		0.000	04:05	
Control Test		0.079	04:05	
Air Blank		0.000	04:05	
Diagnostics Check		OK	04:06	

Data Download Date/Time: 01/29/2016 08:55

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBER,JOSEPH,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: GARCIA,J.R.
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:17

SUBJECT'S LAST NAME: NIGHTINGALE

DATE/TIME OF TEST: 01/29/2016 00:44:37

Diagnostics Check		OK	00:37
Air Blank		0.000	00:38
Control Test		0.080	00:38
Air Blank		0.000	00:38
Subject Sample	# 1	0.180	00:39
Breath Volume		3.437	00:39
Air Blank		0.000	00:40
Air Blank		0.000	00:41
Subject Sample	# 2	0.177	00:42
Breath Volume		2.371	00:42
Air Blank		0.000	00:43
Control Test		0.079	00:43
Air Blank		0.000	00:44
Diagnostics Check		OK	00:44

Data Download Date/Time: 01/29/2016 08:55

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBER,JOSEPH,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BURSE,E
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 01/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: HART

DATE/TIME OF TEST: 01/29/2016 02:08:14

Diagnostics Check		OK	02:01
Air Blank		0.000	02:02
Control Test		0.081	02:02
Air Blank		0.000	02:02
Subject Sample	# 1	0.126	02:03
Breath Volume		1.824	02:03
Air Blank		0.000	02:04
Air Blank		0.000	02:05
Subject Sample	# 2	0.122	02:06
Breath Volume		2.027	02:06
Air Blank		0.000	02:07
Control Test		0.079	02:07
Air Blank		0.000	02:08
Diagnostics Check		OK	02:08

Data Download Date/Time: 01/29/2016 08:55

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBER,JOSEPH,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: KVIES,K.J.
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:03

SUBJECT'S LAST NAME: KING

DATE/TIME OF TEST: 01/29/2016 06:30:08

Diagnostics Check		OK	06:24
Air Blank		0.000	06:24
Control Test		0.080	06:24
Air Blank		0.000	06:25
Subject Sample	# 1	0.000	06:25
Breath Volume		3.210	06:25
Air Blank		0.000	06:26
Air Blank		0.000	06:28
Subject Sample	# 2	0.000	06:28
Breath Volume		3.160	06:28
Air Blank		0.000	06:29
Control Test		0.080	06:29
Air Blank		0.000	06:30
Diagnostics Check		OK	06:30

Data Download Date/Time: 01/29/2016 08:55

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAY,DANIEL,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: GARCIA,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 19:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:36

SUBJECT'S LAST NAME: HARRISON JR

DATE/TIME OF TEST: 01/29/2016 21:04:38

Diagnostics Check		OK	20:57
Air Blank		0.000	20:57
Control Test		0.080	20:58
Air Blank		0.000	20:58
Subject Sample	# 1	0.154	20:59
Breath Volume		1.570	20:59
Air Blank		0.000	21:00
Air Blank		0.000	21:01
Subject Sample	# 2	0.149	21:02
Breath Volume		3.394	21:02
Air Blank		0.000	21:03
Control Test		0.079	21:03
Air Blank		0.000	21:04
Diagnostics Check		OK	21:04

Data Download Date/Time: 02/12/2016 13:53

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JASIUKIEWICZ,MICHAEL,P
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: JASIUKIEWICZ,MICHAEL
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 12/04/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: VILLALOBOS TORRES

DATE/TIME OF TEST: 01/01/2016 03:49:34

Diagnostics Check		OK	03:43
Air Blank		0.000	03:43
Control Test		0.079	03:43
Air Blank		0.000	03:44
Subject Sample	# 1	0.091	03:45
Breath Volume		3.464	03:45
Air Blank		0.000	03:45
Air Blank		0.000	03:47
Subject Sample	# 2	0.091	03:48
Breath Volume		3.167	03:48
Air Blank		0.000	03:48
Control Test		0.080	03:48
Air Blank		0.000	03:49
Diagnostics Check		OK	03:49

Data Download Date/Time: 01/11/2016 11:42

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JASIUKIEWICZ,MICHAEL,P
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: JASIUKIEWICZ,MICHAEL
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 12/04/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: GRAHAM

DATE/TIME OF TEST: 01/02/2016 04:06:26

Diagnostics Check		OK	04:00
Air Blank		0.000	04:00
Control Test		0.078	04:01
Air Blank		0.000	04:01
Subject Sample	# 1	0.074	04:02
Breath Volume		1.828	04:02
Air Blank		0.000	04:02
Air Blank		0.000	04:04
Subject Sample	# 2	0.073	04:05
Breath Volume		1.394	04:05
Air Blank		0.000	04:05
Control Test		0.080	04:05
Air Blank		0.000	04:06
Diagnostics Check		OK	04:06

Data Download Date/Time: 01/11/2016 11:42

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUCHENEY,WESLEY,A
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: SCHIBLER,BRIAN
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 12/04/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: BERCHOU

DATE/TIME OF TEST: 01/03/2016 02:44:45

Diagnostics Check		OK	02:40	
Air Blank		0.000	02:41	
Control Test		0.077	02:41	
Air Blank		0.000	02:42	
Subject Sample	# 1	0.238	02:43	
Breath Volume		1.972	02:43	
Air Blank			02:43	Purge Fail
Air Blank			02:44	Purge Fail

Data Download Date/Time: 01/11/2016 11:42

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUCHENEY,WESLEY,A
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: SCHIBLER,BRIAN
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 12/04/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: BERCHOU

DATE/TIME OF TEST: 01/03/2016 02:59:40

Diagnostics Check		OK	02:53
Air Blank		0.000	02:54
Control Test		0.078	02:54
Air Blank		0.000	02:54
Subject Sample	# 1	0.245	02:55
Breath Volume		1.433	02:55
Air Blank		0.000	02:55
Air Blank		0.000	02:57
Subject Sample	# 2	0.251	02:58
Breath Volume		1.238	02:58
Air Blank		0.000	02:58
Control Test		0.079	02:59
Air Blank		0.000	02:59
Diagnostics Check		OK	02:59

Data Download Date/Time: 01/11/2016 11:42

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WELLS,OFC,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: DRESSMAN,OFC
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 12/04/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: MCMANN

DATE/TIME OF TEST: 01/05/2016 01:52:18

Diagnostics Check		OK	01:45
Air Blank		0.000	01:46
Control Test		0.079	01:46
Air Blank		0.000	01:47
Subject Sample	# 1	0.253	01:47
Breath Volume		4.324	01:47
Air Blank		0.000	01:48
Air Blank		0.000	01:50
Subject Sample	# 2	0.263	01:50
Breath Volume		3.335	01:50
Air Blank		0.000	01:51
Control Test		0.080	01:51
Air Blank		0.000	01:52
Diagnostics Check		OK	01:52

Data Download Date/Time: 01/11/2016 11:42

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHMIDT,RICHARD,W
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: SCHMIDT,RICHARD
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 12/04/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: MCKINNEY

DATE/TIME OF TEST: 01/10/2016 22:44:03

Diagnostics Check		OK	22:40	
Air Blank		0.000	22:41	
Control Test		0.078	22:41	
Air Blank		0.000	22:41	
Subject Sample	# 1		22:42	Subject Test Refused
Breath Volume		0.000	22:42	
Air Blank		0.000	22:43	
Control Test		0.078	22:43	
Air Blank		0.000	22:43	
Diagnostics Check		OK	22:44	

Data Download Date/Time: 01/11/2016 11:42

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001043

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOTT,JOHN,E
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: LOTT,JOHN
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 01/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28814080A3
DRY GAS EXPIRY DATE: 11/05/2016
VIOLATION CODE: ADMIN
ARREST TIME: 17:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:40

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 01/15/2016 20:22:01

Diagnostics Check		OK	20:16
Air Blank		0.000	20:16
Control Test		0.079	20:16
Air Blank		0.000	20:17
Subject Sample	# 1	0.000	20:17
Breath Volume		3.199	20:17
Air Blank		0.000	20:18
Air Blank		0.000	20:20
Subject Sample	# 2	0.000	20:20
Breath Volume		3.093	20:20
Air Blank		0.000	20:21
Control Test		0.079	20:21
Air Blank		0.000	20:21
Diagnostics Check		OK	20:22

Data Download Date/Time: 02/19/2016 13:22

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001043

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOTT,JOHN,E
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: DICK,RICHARD
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 01/15/2016

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 28814080A3
DRY GAS EXPIRY DATE: 11/05/2016
VIOLATION CODE: DUI
ARREST TIME: 20:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:21

SUBJECT'S LAST NAME: WOODS

DATE/TIME OF TEST: 01/15/2016 21:49:50

Diagnostics Check		OK	21:43
Air Blank		0.000	21:43
Control Test		0.081	21:44
Air Blank		0.000	21:44
Subject Sample	# 1	0.179	21:45
Breath Volume		2.410	21:45
Air Blank		0.000	21:45
Air Blank		0.000	21:47
Subject Sample	# 2	0.185	21:48
Breath Volume		2.589	21:48
Air Blank		0.000	21:48
Control Test		0.079	21:49
Air Blank		0.000	21:49
Diagnostics Check		OK	21:49

Data Download Date/Time: 02/19/2016 13:22

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001043

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOTT,JOHN,E
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: RITENOUR,STEVEN
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 01/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28814080A3
DRY GAS EXPIRY DATE: 11/05/2016
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:29

SUBJECT'S LAST NAME: ROWE

DATE/TIME OF TEST: 01/16/2016 02:57:21

Diagnostics Check		OK	02:54	
Air Blank		0.000	02:55	
Control Test		0.080	02:55	
Air Blank		0.000	02:55	
Subject Sample	# 1		02:56	Subject Test Refused
Breath Volume		0.000	02:56	
Air Blank		0.000	02:56	
Control Test		0.080	02:56	
Air Blank		0.000	02:57	
Diagnostics Check		OK	02:57	

Data Download Date/Time: 02/19/2016 13:22

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001043

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORSE,MICHAEL,P
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: PTR,PTR
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 01/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28814080A3
DRY GAS EXPIRY DATE: 11/05/2016
VIOLATION CODE: COURT ORD
ARREST TIME: 09:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:00

SUBJECT'S LAST NAME: SLAUGHTER

DATE/TIME OF TEST: 01/16/2016 09:35:50

Diagnostics Check		OK	09:29
Air Blank		0.000	09:29
Control Test		0.079	09:30
Air Blank		0.000	09:30
Subject Sample	# 1	0.000	09:31
Breath Volume		1.660	09:31
Air Blank		0.000	09:31
Air Blank		0.000	09:33
Subject Sample	# 2	0.000	09:34
Breath Volume		2.847	09:34
Air Blank		0.000	09:34
Control Test		0.080	09:35
Air Blank		0.000	09:35
Diagnostics Check		OK	09:35

Data Download Date/Time: 02/19/2016 13:22

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001043

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAUGHN,DALE,E
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: PTR,PTR
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 01/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28814080A3
DRY GAS EXPIRY DATE: 11/05/2016
VIOLATION CODE: COURT ORD
ARREST TIME: 09:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:00

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 01/17/2016 09:32:22

Diagnostics Check		OK	09:26
Air Blank		0.000	09:26
Control Test		0.081	09:26
Air Blank		0.000	09:27
Subject Sample	# 1	0.000	09:28
Breath Volume		2.535	09:28
Air Blank		0.000	09:28
Air Blank		0.000	09:30
Subject Sample	# 2	0.000	09:31
Breath Volume		2.980	09:31
Air Blank		0.000	09:31
Control Test		0.079	09:31
Air Blank		0.000	09:32
Diagnostics Check		OK	09:32

Data Download Date/Time: 02/19/2016 13:22

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001043

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ABELN,SHON,W
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: ABELN,SHON
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 01/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28814080A3
DRY GAS EXPIRY DATE: 11/05/2016
VIOLATION CODE: COURT ORD
ARREST TIME: 08:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:00

SUBJECT'S LAST NAME: MURRAY

DATE/TIME OF TEST: 01/18/2016 08:46:53

Diagnostics Check		OK	08:41
Air Blank		0.000	08:41
Control Test		0.080	08:41
Air Blank		0.000	08:42
Subject Sample	# 1	0.000	08:42
Breath Volume		1.566	08:42
Air Blank		0.000	08:43
Air Blank		0.000	08:45
Subject Sample	# 2	0.000	08:45
Breath Volume		1.855	08:45
Air Blank		0.000	08:46
Control Test		0.080	08:46
Air Blank		0.000	08:46
Diagnostics Check		OK	08:46

Data Download Date/Time: 02/19/2016 13:22

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001043

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ABELN,SHON,W
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: ABELN,SHON
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 01/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28814080A3
DRY GAS EXPIRY DATE: 11/05/2016
VIOLATION CODE: COURT ORD
ARREST TIME: 16:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:45

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 01/18/2016 17:25:03

Diagnostics Check		OK	17:19
Air Blank		0.000	17:19
Control Test		0.080	17:20
Air Blank		0.000	17:20
Subject Sample	# 1	0.000	17:20
Breath Volume		1.378	17:20
Air Blank		0.000	17:21
Air Blank		0.000	17:23
Subject Sample	# 2	0.000	17:23
Breath Volume		1.437	17:23
Air Blank		0.000	17:24
Control Test		0.080	17:24
Air Blank		0.000	17:24
Diagnostics Check		OK	17:25

Data Download Date/Time: 02/19/2016 13:22

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001043

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOTT,JOHN,E
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: LOTT,JOHN
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 01/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28814080A3
DRY GAS EXPIRY DATE: 11/05/2016
VIOLATION CODE: ADMIN
ARREST TIME: 11:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 11:20

SUBJECT'S LAST NAME: REIMER

DATE/TIME OF TEST: 01/21/2016 11:52:23

Diagnostics Check		OK	11:44
Air Blank		0.000	11:45
Control Test		0.079	11:45
Air Blank		0.000	11:45
Subject Sample	# 1	0.000	11:46
Breath Volume		1.250	11:46
Air Blank		0.000	11:46
Air Blank		0.000	11:48
Subject Sample	# 2	0.000	11:51
Breath Volume		1.343	11:51
Air Blank		0.000	11:51
Control Test		0.080	11:51
Air Blank		0.000	11:52
Diagnostics Check		OK	11:52

Data Download Date/Time: 02/19/2016 13:22

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESCOBAR,ALEJANDRO,
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf
ARRESTING OFFICER: FEISTHAMMEL,SEAN
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf
LAST AGENCY INSPECTION: 12/22/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1698962
DRY GAS EXPIRY DATE: 06/22/2017
VIOLATION CODE: DUI
ARREST TIME: 18:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:42

SUBJECT'S LAST NAME: SCOTTALINE

DATE/TIME OF TEST: 01/16/2016 19:22:30

Diagnostics Check		OK	19:15
Air Blank		0.000	19:16
Control Test		0.079	19:16
Air Blank		0.000	19:17
Subject Sample	# 1	0.000	19:17
Breath Volume		1.855	19:17
Air Blank		0.000	19:18
Air Blank		0.000	19:20
Subject Sample	# 2	0.000	19:21
Breath Volume		1.929	19:21
Air Blank		0.000	19:21
Control Test		0.079	19:21
Air Blank		0.000	19:22
Diagnostics Check		OK	19:22

Data Download Date/Time: 01/26/2016 13:04

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESCOBAR,ALEJANDRO,
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf
ARRESTING OFFICER: ESCOBAR,ALEJANDRO
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf
LAST AGENCY INSPECTION: 12/22/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1698962
DRY GAS EXPIRY DATE: 06/22/2017
VIOLATION CODE: DUI
ARREST TIME: 03:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:17

SUBJECT'S LAST NAME: SOARES DE SA

DATE/TIME OF TEST: 01/17/2016 03:58:05

Diagnostics Check		OK	03:51
Air Blank		0.000	03:51
Control Test		0.078	03:51
Air Blank		0.000	03:52
Subject Sample	# 1	0.162	03:53
Breath Volume		1.382	03:53
Air Blank		0.000	03:54
Air Blank		0.000	03:55
Subject Sample	# 2	0.152	03:56
Breath Volume		1.144	03:56
Air Blank		0.000	03:57
Control Test		0.077	03:57
Air Blank		0.000	03:57
Diagnostics Check		OK	03:58

Data Download Date/Time: 01/26/2016 13:04

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESCOBAR,ALEJANDRO,
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf
ARRESTING OFFICER: ESCOBAR,ALEJANDRO
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf
LAST AGENCY INSPECTION: 12/22/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1698962
DRY GAS EXPIRY DATE: 06/22/2017
VIOLATION CODE: DUI
ARREST TIME: 02:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: ROBINSON

DATE/TIME OF TEST: 01/18/2016 03:37:15

Diagnostics Check		OK	03:34	
Air Blank		0.000	03:34	
Control Test		0.078	03:35	
Air Blank		0.000	03:35	
Subject Sample	# 1		03:35	Subject Test Refused
Breath Volume		0.000	03:35	
Air Blank		0.000	03:36	
Control Test		0.079	03:36	
Air Blank		0.000	03:37	
Diagnostics Check		OK	03:37	

Data Download Date/Time: 01/26/2016 13:04

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESCOBAR,ALEJANDRO,
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf
ARRESTING OFFICER: ESCOBAR,ALEJANDRO
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf
LAST AGENCY INSPECTION: 12/22/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1698962
DRY GAS EXPIRY DATE: 06/22/2017
VIOLATION CODE: DUI
ARREST TIME: 01:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:11

SUBJECT'S LAST NAME: LINEMAN

DATE/TIME OF TEST: 01/22/2016 01:49:58

Diagnostics Check		OK	01:43
Air Blank		0.000	01:44
Control Test		0.079	01:44
Air Blank		0.000	01:44
Subject Sample	# 1	0.126	01:45
Breath Volume		1.460	01:45
Air Blank		0.000	01:45
Air Blank		0.000	01:47
Subject Sample	# 2	0.121	01:48
Breath Volume		1.957	01:48
Air Blank		0.000	01:49
Control Test		0.079	01:49
Air Blank		0.000	01:49
Diagnostics Check		OK	01:49

Data Download Date/Time: 01/26/2016 13:04

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESCOBAR,ALEJANDRO,
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf
ARRESTING OFFICER: ESCOBAR,ALEJANDRO
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf
LAST AGENCY INSPECTION: 12/22/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1698962
DRY GAS EXPIRY DATE: 06/22/2017
VIOLATION CODE: DUI
ARREST TIME: 23:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:53

SUBJECT'S LAST NAME: COVIN

DATE/TIME OF TEST: 01/23/2016 00:30:56

Diagnostics Check		OK	00:24
Air Blank		0.000	00:24
Control Test		0.079	00:25
Air Blank		0.000	00:25
Subject Sample	# 1	0.258	00:26
Breath Volume		1.265	00:26
Air Blank		0.000	00:27
Air Blank		0.000	00:28
Subject Sample	# 2	0.253	00:29
Breath Volume		1.769	00:29
Air Blank		0.000	00:30
Control Test		0.077	00:30
Air Blank		0.000	00:30
Diagnostics Check		OK	00:30

Data Download Date/Time: 01/26/2016 13:04

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESCOBAR,ALEJANDRO,
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf
ARRESTING OFFICER: ESCOBAR,ALEJANDRO
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf
LAST AGENCY INSPECTION: 01/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1698962
DRY GAS EXPIRY DATE: 06/22/2017
VIOLATION CODE: DUI
ARREST TIME: 06:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:40

SUBJECT'S LAST NAME: HELLMERS

DATE/TIME OF TEST: 01/27/2016 07:45:06

Diagnostics Check		OK	07:42	
Air Blank		0.000	07:42	
Control Test		0.079	07:42	
Air Blank		0.000	07:43	
Subject Sample	# 1		07:43	Subject Test Refused
Breath Volume		0.000	07:43	
Air Blank		0.000	07:44	
Control Test		0.079	07:44	
Air Blank		0.000	07:45	
Diagnostics Check		OK	07:45	

Data Download Date/Time: 02/25/2016 12:16

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESCOBAR,ALEJANDRO,
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf
ARRESTING OFFICER: ESCOBAR,ALEJANDRO
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf
LAST AGENCY INSPECTION: 01/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1698962
DRY GAS EXPIRY DATE: 06/22/2017
VIOLATION CODE: DUI
ARREST TIME: 02:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:28

SUBJECT'S LAST NAME: LEON

DATE/TIME OF TEST: 01/28/2016 03:12:05

Diagnostics Check		OK	03:01	
Air Blank		0.000	03:01	
Control Test		0.079	03:01	
Air Blank		0.000	03:02	
Subject Sample	# 1	0.133	03:05	Volume Not Met
Breath Volume		0.867	03:05	
Air Blank		0.000	03:06	
Air Blank		0.000	03:08	
Subject Sample	# 2		03:10	Subject Test Refused
Breath Volume		0.367	03:10	
Air Blank		0.000	03:11	
Control Test		0.078	03:11	
Air Blank		0.000	03:12	
Diagnostics Check		OK	03:12	

Data Download Date/Time: 02/25/2016 12:16

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESCOBAR,ALEJANDRO,
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf
ARRESTING OFFICER: ESCOBAR,ALEJANDRO
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf
LAST AGENCY INSPECTION: 01/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1698962
DRY GAS EXPIRY DATE: 06/22/2017
VIOLATION CODE: DUI
ARREST TIME: 03:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: NOLAN

DATE/TIME OF TEST: 01/31/2016 04:47:14

Diagnostics Check		OK	04:38	
Air Blank		0.000	04:39	
Control Test		0.079	04:39	
Air Blank		0.000	04:40	
Subject Sample	# 1	0.155	04:43	Volume Not Met
Breath Volume		0.359	04:43	
Air Blank		0.000	04:44	
Air Blank		0.000	04:45	
Subject Sample	# 2		04:45	Subject Test Refused
Breath Volume		0.000	04:45	
Air Blank		0.000	04:46	
Control Test		0.077	04:46	
Air Blank		0.000	04:47	
Diagnostics Check		OK	04:47	

Data Download Date/Time: 02/25/2016 12:16

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SAPP,JOSHUA
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 01/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG402702
DRY GAS EXPIRY DATE: 01/27/2016
VIOLATION CODE: DUI
ARREST TIME: 23:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: JORDAN

DATE/TIME OF TEST: 01/30/2016 00:40:35

Diagnostics Check		OK	00:34
Air Blank		0.000	00:34
Control Test		0.081	00:35
Air Blank		0.000	00:35
Subject Sample	# 1	0.068	00:36
Breath Volume		3.265	00:36
Air Blank		0.000	00:36
Air Blank		0.000	00:38
Subject Sample	# 2	0.068	00:39
Breath Volume		3.210	00:39
Air Blank		0.000	00:39
Control Test		0.081	00:40
Air Blank		0.000	00:40
Diagnostics Check		OK	00:40

Data Download Date/Time: 02/25/2016 17:22

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUN,TONY,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ICOBELLI,RICHARD
ARRESTING OFFICER AGENCY: Hallandale Beach PD
LAST AGENCY INSPECTION: 01/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG402702
DRY GAS EXPIRY DATE: 01/27/2016
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:59

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 01/30/2016 01:29:20

Diagnostics Check		OK	01:22
Air Blank		0.000	01:23
Control Test		0.081	01:23
Air Blank		0.000	01:24
Subject Sample	# 1	0.330	01:24
Breath Volume		1.148	01:24
Air Blank		0.000	01:25
Air Blank		0.000	01:27
Subject Sample	# 2	0.332	01:27
Breath Volume		1.984	01:27
Air Blank		0.000	01:28
Control Test		0.081	01:28
Air Blank		0.000	01:29
Diagnostics Check		OK	01:29

Data Download Date/Time: 02/25/2016 17:22

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SEPOT,MICHAEL
ARRESTING OFFICER AGENCY: Margate PD
LAST AGENCY INSPECTION: 01/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG402702
DRY GAS EXPIRY DATE: 01/27/2016
VIOLATION CODE: DUI
ARREST TIME: 03:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: SMALLS

DATE/TIME OF TEST: 01/30/2016 04:14:14

Diagnostics Check		OK	04:07
Air Blank		0.000	04:08
Control Test		0.081	04:08
Air Blank		0.000	04:09
Subject Sample	# 1	0.149	04:09
Breath Volume		1.894	04:09
Air Blank		0.000	04:10
Air Blank		0.000	04:12
Subject Sample	# 2	0.149	04:12
Breath Volume		1.421	04:12
Air Blank		0.000	04:13
Control Test		0.081	04:13
Air Blank		0.000	04:14
Diagnostics Check		OK	04:14

Data Download Date/Time: 02/25/2016 17:22

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINA,MARMILLY,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: MEYERS,ERIC
ARRESTING OFFICER AGENCY: Sunrise PD
LAST AGENCY INSPECTION: 01/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 01/31/2016 02:50:09

Diagnostics Check		OK	02:44
Air Blank		0.000	02:44
Control Test		0.081	02:44
Air Blank		0.000	02:45
Subject Sample	# 1	0.031	02:45
Breath Volume		2.261	02:45
Air Blank		0.000	02:46
Air Blank		0.000	02:48
Subject Sample	# 2	0.031	02:48
Breath Volume		2.441	02:48
Air Blank		0.000	02:49
Control Test		0.081	02:49
Air Blank		0.000	02:50
Diagnostics Check		OK	02:50

Data Download Date/Time: 02/25/2016 17:22

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SERPHOS,SEAN
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 01/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 03:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:17

SUBJECT'S LAST NAME: GUIRGUIS

DATE/TIME OF TEST: 01/31/2016 03:52:02

Diagnostics Check		OK	03:45
Air Blank		0.000	03:45
Control Test		0.081	03:45
Air Blank		0.000	03:46
Subject Sample	# 1	0.162	03:47
Breath Volume		1.875	03:47
Air Blank		0.000	03:47
Air Blank		0.000	03:49
Subject Sample	# 2	0.151	03:50
Breath Volume		1.246	03:50
Air Blank		0.000	03:51
Control Test		0.081	03:51
Air Blank		0.000	03:51
Diagnostics Check		OK	03:52

Data Download Date/Time: 02/25/2016 17:22

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CINELLI,ADAM,D
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: CINELLI,ADAM
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 12/02/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A1
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:04

SUBJECT'S LAST NAME: ALGIRE

DATE/TIME OF TEST: 01/01/2016 02:31:22

Diagnostics Check		OK	02:25
Air Blank		0.000	02:25
Control Test		0.079	02:26
Air Blank		0.000	02:26
Subject Sample	# 1	0.000	02:26
Breath Volume		1.375	02:26
Air Blank		0.000	02:27
Air Blank		0.000	02:29
Subject Sample	# 2	0.000	02:30
Breath Volume		1.898	02:30
Air Blank		0.000	02:30
Control Test		0.080	02:30
Air Blank		0.000	02:31
Diagnostics Check		OK	02:31

Data Download Date/Time: 01/19/2016 13:32

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: IVEY,BRIAN,J
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: IVEY,BRIAN
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 12/02/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A1
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: STRADONE

DATE/TIME OF TEST: 01/01/2016 03:12:55

Diagnostics Check		OK	03:06
Air Blank		0.000	03:07
Control Test		0.080	03:07
Air Blank		0.000	03:07
Subject Sample	# 1	0.082	03:08
Breath Volume		2.027	03:08
Air Blank		0.000	03:09
Air Blank		0.000	03:10
Subject Sample	# 2	0.082	03:11
Breath Volume		1.679	03:11
Air Blank		0.000	03:12
Control Test		0.080	03:12
Air Blank		0.000	03:12
Diagnostics Check		OK	03:12

Data Download Date/Time: 01/19/2016 13:32

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: IVEY,BRIAN,J
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: DAUGHERTY,DANIEL
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 12/02/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A1
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:09

SUBJECT'S LAST NAME: HALL

DATE/TIME OF TEST: 01/03/2016 00:35:56

Diagnostics Check		OK	00:29
Air Blank		0.000	00:29
Control Test		0.080	00:30
Air Blank		0.000	00:30
Subject Sample	# 1	0.208	00:31
Breath Volume		2.421	00:31
Air Blank		0.000	00:32
Air Blank		0.000	00:33
Subject Sample	# 2	0.197	00:34
Breath Volume		2.640	00:34
Air Blank		0.000	00:35
Control Test		0.079	00:35
Air Blank		0.000	00:35
Diagnostics Check		OK	00:35

Data Download Date/Time: 01/19/2016 13:32

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHUCK,SCOTT,M
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: MORRISON,ARTHUR
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 12/02/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A1
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:52

SUBJECT'S LAST NAME: GILMORE

DATE/TIME OF TEST: 01/05/2016 01:18:31

Diagnostics Check		OK	01:13	
Air Blank		0.000	01:13	
Control Test		0.080	01:13	
Air Blank		0.000	01:14	
Subject Sample	# 1		01:17	Subject Test Refused
Breath Volume		0.000	01:17	
Air Blank		0.000	01:17	
Control Test		0.081	01:17	
Air Blank		0.000	01:18	
Diagnostics Check		OK	01:18	

Data Download Date/Time: 01/19/2016 13:32

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KELLY,WESLEY,R
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: KELLY,WESLEY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/02/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A1
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: HENNESSY

DATE/TIME OF TEST: 01/06/2016 02:57:21

Diagnostics Check		OK	02:50
Air Blank		0.000	02:51
Control Test		0.080	02:51
Air Blank		0.000	02:51
Subject Sample	# 1	0.186	02:52
Breath Volume		1.945	02:52
Air Blank		0.000	02:53
Air Blank		0.000	02:55
Subject Sample	# 2	0.189	02:55
Breath Volume		2.265	02:55
Air Blank		0.000	02:56
Control Test		0.080	02:56
Air Blank		0.000	02:57
Diagnostics Check		OK	02:57

Data Download Date/Time: 01/19/2016 13:32

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KELLY,WESLEY,R
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: KELLY,WESLEY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/02/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A1
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: SETTY

DATE/TIME OF TEST: 01/07/2016 02:23:24

Diagnostics Check		OK	02:20	
Air Blank		0.000	02:21	
Control Test		0.080	02:21	
Air Blank		0.000	02:21	
Subject Sample	# 1		02:22	Subject Test Refused
Breath Volume		0.000	02:22	
Air Blank		0.000	02:22	
Control Test		0.079	02:22	
Air Blank		0.000	02:23	
Diagnostics Check		OK	02:23	

Data Download Date/Time: 01/19/2016 13:32

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,R,J
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: EVANS,R
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/02/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A1
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: WALDEN

DATE/TIME OF TEST: 01/07/2016 04:18:18

Diagnostics Check		OK	04:06	
Air Blank		0.000	04:07	
Control Test		0.079	04:07	
Air Blank		0.000	04:08	
Subject Sample	# 1		04:11	No Sample Provided
Breath Volume		0.000	04:11	
Air Blank		0.000	04:11	
Air Blank		0.000	04:13	
Subject Sample	# 2		04:16	No Sample Provided
Breath Volume		0.000	04:16	
Air Blank		0.000	04:17	
Control Test		0.080	04:17	
Air Blank		0.000	04:18	
Diagnostics Check		OK	04:18	

Data Download Date/Time: 01/19/2016 13:32

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NOTTINGHAM,CHRISTOPHER,R
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: NOTTINGHAM,CHRISTOPHER
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/02/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A1
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: MCRAE

DATE/TIME OF TEST: 01/08/2016 22:27:38

Diagnostics Check		OK	22:20
Air Blank		0.000	22:21
Control Test		0.081	22:21
Air Blank		0.000	22:22
Subject Sample	# 1	0.037	22:23
Breath Volume		1.910	22:23
Air Blank		0.000	22:23
Air Blank		0.000	22:25
Subject Sample	# 2	0.035	22:26
Breath Volume		3.093	22:26
Air Blank		0.000	22:26
Control Test		0.081	22:27
Air Blank		0.000	22:27
Diagnostics Check		OK	22:27

Data Download Date/Time: 01/19/2016 13:32

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOLAN,NICHOLAS,C
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: DOLAN,NICHOLAS
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/02/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A1
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:03

SUBJECT'S LAST NAME: WISDOM

DATE/TIME OF TEST: 01/09/2016 03:33:22

Diagnostics Check		OK	03:27
Air Blank		0.000	03:27
Control Test		0.080	03:28
Air Blank		0.000	03:28
Subject Sample	# 1	0.009	03:29
Breath Volume		1.652	03:29
Air Blank		0.000	03:29
Air Blank		0.000	03:31
Subject Sample	# 2	0.008	03:32
Breath Volume		1.714	03:32
Air Blank		0.000	03:32
Control Test		0.081	03:32
Air Blank		0.000	03:33
Diagnostics Check		OK	03:33

Data Download Date/Time: 01/19/2016 13:32

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANGELICCHI,JOSEPH,S
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: ANGELICCHI,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/02/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A1
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:43

SUBJECT'S LAST NAME: MAGIN

DATE/TIME OF TEST: 01/09/2016 05:09:49

Diagnostics Check		OK	05:03
Air Blank		0.000	05:04
Control Test		0.079	05:04
Air Blank		0.000	05:04
Subject Sample	# 1	0.148	05:05
Breath Volume		1.320	05:05
Air Blank		0.000	05:05
Air Blank		0.000	05:07
Subject Sample	# 2	0.144	05:08
Breath Volume		1.906	05:08
Air Blank		0.000	05:08
Control Test		0.080	05:09
Air Blank		0.000	05:09
Diagnostics Check		OK	05:09

Data Download Date/Time: 01/19/2016 13:32

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANGELICCHI,JOSEPH,S
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: CORBINO,DANIEL
ARRESTING OFFICER AGENCY: New Port Richey PD
LAST AGENCY INSPECTION: 12/02/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A1
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:05

SUBJECT'S LAST NAME: GRAY

DATE/TIME OF TEST: 01/09/2016 05:31:57

Diagnostics Check		OK	05:25
Air Blank		0.000	05:26
Control Test		0.078	05:26
Air Blank		0.000	05:26
Subject Sample	# 1	0.124	05:27
Breath Volume		2.476	05:27
Air Blank		0.000	05:28
Air Blank		0.000	05:30
Subject Sample	# 2	0.128	05:30
Breath Volume		2.433	05:30
Air Blank		0.000	05:31
Control Test		0.079	05:31
Air Blank		0.000	05:31
Diagnostics Check		OK	05:31

Data Download Date/Time: 01/19/2016 13:32

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NOTTINGHAM,CHRISTOPHER,R
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: NOTTINGHAM,CHRISTOPHER
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/02/2015

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 03415080A1
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 07:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:19

SUBJECT'S LAST NAME: METZ

DATE/TIME OF TEST: 01/09/2016 21:43:08

Diagnostics Check		OK	21:39	
Air Blank		0.000	21:40	
Control Test		0.080	21:40	
Air Blank		0.000	21:40	
Subject Sample	# 1		21:41	Subject Test Refused
Breath Volume		0.000	21:41	
Air Blank		0.000	21:42	
Control Test		0.080	21:42	
Air Blank		0.000	21:43	
Diagnostics Check		OK	21:43	

Data Download Date/Time: 01/19/2016 13:32

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KELLY,WESLEY,R
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: BLICKER,EVAN
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 12/02/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A1
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:01

SUBJECT'S LAST NAME: CROCKETT

DATE/TIME OF TEST: 01/10/2016 05:58:35

Diagnostics Check		OK	05:51
Air Blank		0.000	05:51
Control Test		0.079	05:52
Air Blank		0.000	05:52
Subject Sample	# 1	0.160	05:53
Breath Volume		2.160	05:53
Air Blank		0.000	05:54
Air Blank		0.000	05:56
Subject Sample	# 2	0.170	05:57
Breath Volume		2.519	05:57
Air Blank		0.000	05:57
Control Test		0.080	05:58
Air Blank		0.000	05:58
Diagnostics Check		OK	05:58

Data Download Date/Time: 01/19/2016 13:32

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOLAN,NICHOLAS,C
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: DOLAN,NICHOLAS
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/02/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A1
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 19:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:26

SUBJECT'S LAST NAME: JACOBY

DATE/TIME OF TEST: 01/10/2016 19:52:27

Diagnostics Check		OK	19:46
Air Blank		0.000	19:46
Control Test		0.081	19:46
Air Blank		0.000	19:47
Subject Sample	# 1	0.134	19:47
Breath Volume		4.164	19:47
Air Blank		0.000	19:48
Air Blank		0.000	19:50
Subject Sample	# 2	0.133	19:50
Breath Volume		4.359	19:50
Air Blank		0.000	19:51
Control Test		0.080	19:51
Air Blank		0.000	19:52
Diagnostics Check		OK	19:52

Data Download Date/Time: 01/19/2016 13:32

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KELLY,WESLEY,R
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: KELLY,WESLEY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/02/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A1
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: HIDALGO ESTRADA

DATE/TIME OF TEST: 01/11/2016 00:07:41

Diagnostics Check		OK	00:01
Air Blank		0.000	00:01
Control Test		0.080	00:01
Air Blank		0.000	00:02
Subject Sample	# 1	0.188	00:03
Breath Volume		1.906	00:03
Air Blank		0.000	00:03
Air Blank		0.000	00:05
Subject Sample	# 2	0.190	00:06
Breath Volume		2.402	00:06
Air Blank		0.000	00:06
Control Test		0.080	00:07
Air Blank		0.000	00:07
Diagnostics Check		OK	00:07

Data Download Date/Time: 01/19/2016 13:32

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KELLY,WESLEY,R
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: KELLY,WESLEY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/02/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A1
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 11:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: GRECO

DATE/TIME OF TEST: 01/12/2016 00:58:49

Diagnostics Check		OK	00:51
Air Blank		0.000	00:52
Control Test		0.082	00:52
Air Blank		0.000	00:53
Subject Sample	# 1	0.083	00:53
Breath Volume		1.648	00:53
Air Blank		0.000	00:54
Air Blank		0.000	00:56
Subject Sample	# 2	0.081	00:57
Breath Volume		2.101	00:57
Air Blank		0.000	00:57
Control Test		0.081	00:58
Air Blank		0.000	00:58
Diagnostics Check		OK	00:58

Data Download Date/Time: 01/19/2016 13:32

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOLAN,NICHOLAS,C
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: DOLAN,NICHOLAS
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/02/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A1
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:28

SUBJECT'S LAST NAME: HOLLAND

DATE/TIME OF TEST: 01/12/2016 01:22:21

Diagnostics Check		OK	01:16	
Air Blank		0.000	01:16	
Control Test		0.080	01:16	
Air Blank		0.000	01:17	
Subject Sample	# 1	0.141	01:18	
Breath Volume		2.089	01:18	
Air Blank		0.000	01:18	
Air Blank		0.000	01:20	
Subject Sample	# 2		01:20	Slope Not Met
Breath Volume		2.101	01:20	
Air Blank		0.000	01:21	
Control Test		0.080	01:21	
Air Blank		0.000	01:22	
Diagnostics Check		OK	01:22	

Data Download Date/Time: 01/19/2016 13:32

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOLAN,NICHOLAS,C
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: DOLAN,NICHOLAS
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/02/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A1
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:28

SUBJECT'S LAST NAME: HOLLAND

DATE/TIME OF TEST: 01/12/2016 01:27:24

Diagnostics Check		OK	01:24	
Air Blank		0.000	01:24	
Control Test		0.081	01:25	
Air Blank		0.000	01:25	
Subject Sample	# 1		01:25	Slope Not Met
Breath Volume		2.328	01:25	
Air Blank		0.000	01:26	
Control Test		0.081	01:26	
Air Blank		0.000	01:27	
Diagnostics Check		OK	01:27	

Data Download Date/Time: 01/19/2016 13:32

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,WILLIAM,T
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SMITH,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/02/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A1
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:21

SUBJECT'S LAST NAME: RINESMITH

DATE/TIME OF TEST: 01/12/2016 02:48:54

Diagnostics Check		OK	02:42
Air Blank		0.000	02:43
Control Test		0.080	02:43
Air Blank		0.000	02:43
Subject Sample	# 1	0.134	02:44
Breath Volume		2.554	02:44
Air Blank		0.000	02:45
Air Blank		0.000	02:47
Subject Sample	# 2	0.127	02:47
Breath Volume		2.175	02:47
Air Blank		0.000	02:48
Control Test		0.080	02:48
Air Blank		0.000	02:48
Diagnostics Check		OK	02:48

Data Download Date/Time: 01/19/2016 13:32

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KELLY,WESLEY,R
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: KELLY,WESLEY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/02/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A1
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:33

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 01/12/2016 04:59:38

Diagnostics Check		OK	04:53
Air Blank		0.000	04:54
Control Test		0.081	04:54
Air Blank		0.000	04:54
Subject Sample	# 1	0.000	04:55
Breath Volume		2.890	04:55
Air Blank		0.000	04:55
Air Blank		0.000	04:57
Subject Sample	# 2	0.000	04:58
Breath Volume		3.050	04:58
Air Blank		0.000	04:58
Control Test		0.081	04:59
Air Blank		0.000	04:59
Diagnostics Check		OK	04:59

Data Download Date/Time: 01/19/2016 13:32

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KELLY,WESLEY,R
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: KELLY,WESLEY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/02/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A1
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: NIEBEL

DATE/TIME OF TEST: 01/13/2016 23:42:49

Diagnostics Check		OK	23:35
Air Blank		0.000	23:35
Control Test		0.080	23:36
Air Blank		0.000	23:36
Subject Sample	# 1	0.243	23:37
Breath Volume		1.824	23:37
Air Blank		0.000	23:38
Air Blank		0.000	23:40
Subject Sample	# 2	0.238	23:41
Breath Volume		1.707	23:41
Air Blank		0.000	23:41
Control Test		0.081	23:42
Air Blank		0.000	23:42
Diagnostics Check		OK	23:42

Data Download Date/Time: 01/19/2016 13:32

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRIOLAT,CASEY,J
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: BRIOLAT,CASEY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/02/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A1
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:18

SUBJECT'S LAST NAME: JENNINGS

DATE/TIME OF TEST: 01/16/2016 03:47:34

Diagnostics Check		OK	03:41
Air Blank		0.000	03:41
Control Test		0.079	03:41
Air Blank		0.000	03:42
Subject Sample	# 1	0.202	03:43
Breath Volume		2.667	03:43
Air Blank		0.000	03:43
Air Blank		0.000	03:45
Subject Sample	# 2	0.191	03:46
Breath Volume		1.515	03:46
Air Blank		0.000	03:46
Control Test		0.079	03:46
Air Blank		0.000	03:47
Diagnostics Check		OK	03:47

Data Download Date/Time: 01/19/2016 13:32

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOLAN,NICHOLAS,C
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: DOLAN,NICHOLAS
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/02/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A1
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 16:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:50

SUBJECT'S LAST NAME: CHILLURA

DATE/TIME OF TEST: 01/16/2016 17:28:34

Diagnostics Check		OK	17:22
Air Blank		0.000	17:22
Control Test		0.079	17:23
Air Blank		0.000	17:23
Subject Sample	# 1	0.090	17:24
Breath Volume		2.937	17:24
Air Blank		0.000	17:24
Air Blank		0.000	17:26
Subject Sample	# 2	0.091	17:27
Breath Volume		2.347	17:27
Air Blank		0.000	17:27
Control Test		0.079	17:27
Air Blank		0.000	17:28
Diagnostics Check		OK	17:28

Data Download Date/Time: 01/19/2016 13:32

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AIKEN,DURRIEL,R
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: AIKEN,DURRIEL
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 12/02/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A1
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: ALMARAZ-LOPEZ

DATE/TIME OF TEST: 01/16/2016 23:22:15

Diagnostics Check		OK	23:13
Air Blank		0.000	23:14
Control Test		0.080	23:14
Air Blank		0.000	23:14
Subject Sample	# 1	0.131	23:17
Breath Volume		1.410	23:17
Air Blank		0.000	23:18
Air Blank		0.000	23:19
Subject Sample	# 2	0.132	23:20
Breath Volume		1.609	23:20
Air Blank		0.000	23:21
Control Test		0.079	23:21
Air Blank		0.000	23:22
Diagnostics Check		OK	23:22

Data Download Date/Time: 01/19/2016 13:33

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOLAN,NICHOLAS,C
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: DOLAN,NICHOLAS
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/02/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A1
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 17:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:48

SUBJECT'S LAST NAME: EZYK

DATE/TIME OF TEST: 01/17/2016 18:28:19

Diagnostics Check		OK	18:22
Air Blank		0.000	18:22
Control Test		0.080	18:23
Air Blank		0.000	18:23
Subject Sample	# 1	0.007	18:23
Breath Volume		1.601	18:23
Air Blank		0.000	18:24
Air Blank		0.000	18:26
Subject Sample	# 2	0.000	18:26
Breath Volume		1.179	18:26
Air Blank		0.000	18:27
Control Test		0.080	18:27
Air Blank		0.000	18:28
Diagnostics Check		OK	18:28

Data Download Date/Time: 01/19/2016 13:33

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOUTROS,PHILLIP,A
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: BOUTROS,PHILLIP
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/02/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A1
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 19:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:22

SUBJECT'S LAST NAME: ZOOK

DATE/TIME OF TEST: 01/17/2016 20:22:03

Diagnostics Check		OK	20:15
Air Blank		0.000	20:16
Control Test		0.080	20:16
Air Blank		0.000	20:17
Subject Sample	# 1	0.096	20:17
Breath Volume		3.375	20:17
Air Blank		0.000	20:18
Air Blank		0.000	20:20
Subject Sample	# 2	0.098	20:20
Breath Volume		3.699	20:20
Air Blank		0.000	20:21
Control Test		0.081	20:21
Air Blank		0.000	20:21
Diagnostics Check		OK	20:22

Data Download Date/Time: 01/19/2016 13:33

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENBO,A,R
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: DENBO,A
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 12/02/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A1
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 17:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:15

SUBJECT'S LAST NAME: COTHRAN

DATE/TIME OF TEST: 01/18/2016 19:43:25

Diagnostics Check		OK	19:36
Air Blank		0.000	19:36
Control Test		0.081	19:36
Air Blank		0.000	19:37
Subject Sample	# 1	0.211	19:38
Breath Volume		1.304	19:38
Air Blank		0.000	19:39
Air Blank		0.000	19:40
Subject Sample	# 2	0.204	19:41
Breath Volume		1.453	19:41
Air Blank		0.000	19:42
Control Test		0.081	19:42
Air Blank		0.000	19:43
Diagnostics Check		OK	19:43

Data Download Date/Time: 01/19/2016 13:33

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001050

INSTRUMENT REGISTERED TO: CORAL SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WETHERINGTON,STEPHEN,L
BREATH TEST OPERATOR AGENCY: Coral Springs PD
ARRESTING OFFICER: WETHERINGTON,STEPHEN
ARRESTING OFFICER AGENCY: Coral Springs PD
LAST AGENCY INSPECTION: 12/16/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:31

SUBJECT'S LAST NAME: RAMIREZ

DATE/TIME OF TEST: 01/03/2016 05:05:53

Diagnostics Check		OK	04:59
Air Blank		0.000	05:00
Control Test		0.079	05:00
Air Blank		0.000	05:00
Subject Sample	# 1	0.149	05:01
Breath Volume		3.429	05:01
Air Blank		0.000	05:02
Air Blank		0.000	05:03
Subject Sample	# 2	0.152	05:04
Breath Volume		3.871	05:04
Air Blank		0.000	05:04
Control Test		0.080	05:05
Air Blank		0.000	05:05
Diagnostics Check		OK	05:05

Data Download Date/Time: 01/27/2016 15:27

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001050

INSTRUMENT REGISTERED TO: CORAL SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHRISTOPHER,CHRISTOPHER,E
BREATH TEST OPERATOR AGENCY: Coral Springs PD
ARRESTING OFFICER: MURDOCK,TROY
ARRESTING OFFICER AGENCY: Coral Springs PD
LAST AGENCY INSPECTION: 12/16/2015

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 06:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:58

SUBJECT'S LAST NAME: JEROME

DATE/TIME OF TEST: 01/23/2016 07:42:45

Diagnostics Check		OK	07:35
Air Blank		0.000	07:36
Control Test		0.080	07:36
Air Blank		0.000	07:37
Subject Sample	# 1	0.190	07:38
Breath Volume		1.480	07:38
Air Blank		0.000	07:38
Air Blank		0.000	07:40
Subject Sample	# 2	0.194	07:41
Breath Volume		1.566	07:41
Air Blank		0.000	07:41
Control Test		0.079	07:42
Air Blank		0.000	07:42
Diagnostics Check		OK	07:42

Data Download Date/Time: 01/27/2016 15:27

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001050

INSTRUMENT REGISTERED TO: CORAL SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: N,STEPHEN,L
BREATH TEST OPERATOR AGENCY: Coral Springs PD
ARRESTING OFFICER: WETHERINGTON,STEPHEN
ARRESTING OFFICER AGENCY: Coral Springs PD
LAST AGENCY INSPECTION: 01/27/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:26

SUBJECT'S LAST NAME: GARZA

DATE/TIME OF TEST: 01/31/2016 01:08:02

Diagnostics Check		OK	00:55	
Air Blank		0.000	00:55	
Control Test		0.080	00:56	
Air Blank		0.000	00:56	
Subject Sample	# 1	0.173	00:57	
Breath Volume		1.355	00:57	
Air Blank		0.000	00:58	
Air Blank		0.000	01:00	
Subject Sample	# 2	0.170	01:03	Volume Not Met
Breath Volume		0.605	01:03	
Air Blank		0.000	01:04	
Air Blank		0.000	01:06	
Subject Sample	# 3	0.172	01:06	
Breath Volume		2.367	01:06	
Air Blank		0.000	01:07	
Control Test		0.079	01:07	
Air Blank		0.000	01:07	
Diagnostics Check		OK	01:08	

Data Download Date/Time: 02/25/2016 13:58

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZOLLOTA,VLADIMIR,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ALVEY,KYLE
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/25/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 02:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: SCHLATTERER

DATE/TIME OF TEST: 01/01/2016 02:54:58

Diagnostics Check		OK	02:51	
Air Blank		0.000	02:52	
Control Test		0.078	02:52	
Air Blank		0.000	02:52	
Subject Sample	# 1		02:53	Subject Test Refused
Breath Volume		0.000	02:53	
Air Blank		0.000	02:54	
Control Test		0.078	02:54	
Air Blank		0.000	02:54	
Diagnostics Check		OK	02:54	

Data Download Date/Time: 01/30/2016 15:10

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZOLLOTA,VLADIMIR,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ARKOVICH,VLADIMIR
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/25/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 01:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: OWENS

DATE/TIME OF TEST: 01/02/2016 02:21:09

Diagnostics Check		OK	02:18	
Air Blank		0.000	02:18	
Control Test		0.079	02:19	
Air Blank		0.000	02:19	
Subject Sample	# 1		02:19	Subject Test Refused
Breath Volume		0.000	02:19	
Air Blank		0.000	02:20	
Control Test		0.079	02:20	
Air Blank		0.000	02:21	
Diagnostics Check		OK	02:21	

Data Download Date/Time: 01/30/2016 15:10

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTER,MATTHEW,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ALVEY,DOUGLAS
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/25/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 00:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: KUVIK

DATE/TIME OF TEST: 01/03/2016 01:01:49

Diagnostics Check		OK	00:52
Air Blank		0.000	00:52
Control Test		0.078	00:53
Air Blank		0.000	00:53
Subject Sample	# 1	0.145	00:56
Breath Volume		1.484	00:56
Air Blank		0.000	00:57
Air Blank		0.000	00:58
Subject Sample	# 2	0.152	01:00
Breath Volume		1.898	01:00
Air Blank		0.000	01:00
Control Test		0.078	01:01
Air Blank		0.000	01:01
Diagnostics Check		OK	01:01

Data Download Date/Time: 01/30/2016 15:10

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZOLLOTA,VLADIMIR,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CARTER,MATTHEW
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/25/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 01:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: WILTSHIRE

DATE/TIME OF TEST: 01/03/2016 02:20:55

Diagnostics Check		OK	02:17	
Air Blank		0.000	02:17	
Control Test		0.077	02:17	
Air Blank		0.000	02:18	
Subject Sample	# 1		02:19	RFI Detect
Breath Volume		0.000	02:19	
Air Blank		0.000	02:20	
Control Test		0.078	02:20	
Air Blank		0.000	02:20	
Diagnostics Check		OK	02:20	

Data Download Date/Time: 01/30/2016 15:10

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZOLLOTA,VLADIMIR,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CARTER,MATTHEW
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/25/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 01:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: WILTSHIRE

DATE/TIME OF TEST: 01/03/2016 02:25:23

Diagnostics Check		OK	02:22	
Air Blank		0.000	02:22	
Control Test		0.077	02:22	
Air Blank		0.000	02:23	
Subject Sample	# 1		02:23	Slope Not Met
Breath Volume		1.191	02:23	
Air Blank		0.000	02:24	
Control Test		0.077	02:24	
Air Blank		0.000	02:25	
Diagnostics Check		OK	02:25	

Data Download Date/Time: 01/30/2016 15:10

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZOLLOTA,VLADIMIR,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CARTER,MATTHEW
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/25/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 01:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: WILTSHIRE

DATE/TIME OF TEST: 01/03/2016 02:34:28

Diagnostics Check		OK	02:26
Air Blank		0.000	02:27
Control Test		0.077	02:27
Air Blank		0.000	02:28
Subject Sample	# 1	0.226	02:28
Breath Volume		2.281	02:28
Air Blank		0.000	02:29
Air Blank		0.000	02:31
Subject Sample	# 2	0.234	02:32
Breath Volume		2.316	02:32
Air Blank		0.000	02:33
Control Test		0.077	02:33
Air Blank		0.000	02:34
Diagnostics Check		OK	02:34

Data Download Date/Time: 01/30/2016 15:10

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTER,MATTHEW,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ZOLLOTA,VLADIMIR
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/25/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 03:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: LEWIS

DATE/TIME OF TEST: 01/03/2016 03:23:58

Diagnostics Check		OK	03:17
Air Blank		0.000	03:17
Control Test		0.078	03:17
Air Blank		0.000	03:18
Subject Sample	# 1	0.279	03:19
Breath Volume		1.562	03:19
Air Blank		0.000	03:19
Air Blank		0.000	03:21
Subject Sample	# 2	0.296	03:22
Breath Volume		1.886	03:22
Air Blank		0.000	03:23
Control Test		0.077	03:23
Air Blank		0.000	03:23
Diagnostics Check		OK	03:23

Data Download Date/Time: 01/30/2016 15:10

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCE JR,SCOTT,A
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: PIERCE JR,SCOTT
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/25/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 11:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:25

SUBJECT'S LAST NAME: SANDEERS

DATE/TIME OF TEST: 01/06/2016 12:06:37

Diagnostics Check		OK	11:57
Air Blank		0.000	11:57
Control Test		0.080	11:58
Air Blank		0.000	11:58
Subject Sample	# 1	0.293	11:59
Breath Volume		1.843	11:59
Air Blank		0.000	11:59
Air Blank		0.000	12:01
Subject Sample	# 2	0.265	12:02
Breath Volume		1.351	12:02
Air Blank		0.000	12:02
Air Blank		0.000	12:04
Subject Sample	# 3	0.270	12:05
Breath Volume		1.843	12:05
Air Blank		0.000	12:05
Control Test		0.078	12:06
Air Blank		0.000	12:06
Diagnostics Check		OK	12:06

Data Download Date/Time: 01/30/2016 15:10

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALVEY,DOUGLAS,J
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ALVEY,DOUGLAS
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/25/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: SPARKS

DATE/TIME OF TEST: 01/07/2016 01:15:29

Diagnostics Check		OK	01:12	
Air Blank		0.000	01:13	
Control Test		0.079	01:13	
Air Blank		0.000	01:13	
Subject Sample	# 1		01:14	Subject Test Refused
Breath Volume		0.000	01:14	
Air Blank		0.000	01:14	
Control Test		0.077	01:14	
Air Blank		0.000	01:15	
Diagnostics Check		OK	01:15	

Data Download Date/Time: 01/30/2016 15:10

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KRONZ,BRIAN,M
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: KRONZ,BRIAN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/25/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 18:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:55

SUBJECT'S LAST NAME: MINOR

DATE/TIME OF TEST: 01/08/2016 19:22:16

Diagnostics Check		OK	19:16
Air Blank		0.000	19:16
Control Test		0.078	19:16
Air Blank		0.000	19:17
Subject Sample	# 1	0.174	19:17
Breath Volume		2.382	19:17
Air Blank		0.000	19:18
Air Blank		0.000	19:20
Subject Sample	# 2	0.177	19:20
Breath Volume		2.425	19:20
Air Blank		0.000	19:21
Control Test		0.077	19:21
Air Blank		0.000	19:22
Diagnostics Check		OK	19:22

Data Download Date/Time: 01/30/2016 15:10

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTER,MATTHEW,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ZOLLOTA,VLADIMIR
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/25/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 01:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:56

SUBJECT'S LAST NAME: SILERS

DATE/TIME OF TEST: 01/10/2016 02:20:12

Diagnostics Check		OK	02:16	
Air Blank		0.000	02:16	
Control Test		0.078	02:17	
Air Blank		0.000	02:17	
Subject Sample	# 1		02:18	Subject Test Refused
Breath Volume		0.000	02:18	
Air Blank		0.000	02:19	
Control Test		0.078	02:19	
Air Blank		0.000	02:20	
Diagnostics Check		OK	02:20	

Data Download Date/Time: 01/30/2016 15:10

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZOLLOTA,VLADIMIR,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CARTER,MATTHEW
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/25/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 03:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:37

SUBJECT'S LAST NAME: JAMES

DATE/TIME OF TEST: 01/14/2016 04:04:51

Diagnostics Check		OK	04:01	
Air Blank		0.000	04:01	
Control Test		0.078	04:02	
Air Blank		0.000	04:02	
Subject Sample	# 1		04:03	Subject Test Refused
Breath Volume		0.000	04:03	
Air Blank		0.000	04:03	
Control Test		0.078	04:04	
Air Blank		0.000	04:04	
Diagnostics Check		OK	04:04	

Data Download Date/Time: 01/30/2016 15:10

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALVEY,DOUGLAS,J
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ZOLLOTA,VLADAMIR
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/25/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 01:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: STEUSSY-WILLIAMS

DATE/TIME OF TEST: 01/17/2016 02:37:16

Diagnostics Check		OK	02:30
Air Blank		0.000	02:31
Control Test		0.079	02:31
Air Blank		0.000	02:32
Subject Sample	# 1	0.230	02:32
Breath Volume		2.230	02:32
Air Blank		0.000	02:33
Air Blank		0.000	02:35
Subject Sample	# 2	0.220	02:35
Breath Volume		1.667	02:35
Air Blank		0.000	02:36
Control Test		0.078	02:36
Air Blank		0.000	02:37
Diagnostics Check		OK	02:37

Data Download Date/Time: 01/30/2016 15:10

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HASSETT,STEVEN,J
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: HASSETT,STEVEN
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 12/25/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 22:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 01/18/2016 00:28:23

Diagnostics Check		OK	00:25	
Air Blank		0.000	00:25	
Control Test		0.078	00:26	
Air Blank		0.000	00:26	
Subject Sample	# 1		00:27	Subject Test Refused
Breath Volume		0.000	00:27	
Air Blank		0.000	00:27	
Control Test		0.079	00:27	
Air Blank		0.000	00:28	
Diagnostics Check		OK	00:28	

Data Download Date/Time: 01/30/2016 15:10

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTER,MATTHEW,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CARTER,MATTHEW
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/25/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 20:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:15

SUBJECT'S LAST NAME: LUGO

DATE/TIME OF TEST: 01/22/2016 21:45:44

Diagnostics Check		OK	21:38
Air Blank		0.000	21:38
Control Test		0.079	21:39
Air Blank		0.000	21:39
Subject Sample	# 1	0.221	21:40
Breath Volume		1.460	21:40
Air Blank		0.000	21:41
Air Blank		0.000	21:43
Subject Sample	# 2	0.222	21:44
Breath Volume		1.460	21:44
Air Blank		0.000	21:44
Control Test		0.078	21:45
Air Blank		0.000	21:45
Diagnostics Check		OK	21:45

Data Download Date/Time: 01/30/2016 15:10

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTER,MATTHEW,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ZOLLOTA,VLADIMIR
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/25/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 02:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 01/24/2016 02:41:04

Diagnostics Check		OK	02:34
Air Blank		0.000	02:35
Control Test		0.079	02:35
Air Blank		0.000	02:35
Subject Sample	# 1	0.000	02:36
Breath Volume		2.132	02:36
Air Blank		0.000	02:37
Air Blank		0.000	02:39
Subject Sample	# 2	0.000	02:39
Breath Volume		1.464	02:39
Air Blank		0.000	02:40
Control Test		0.079	02:40
Air Blank		0.000	02:40
Diagnostics Check		OK	02:41

Data Download Date/Time: 01/30/2016 15:10

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KRONZ,BRIAN,M
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: KRONZ,BRIAN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/25/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 18:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:45

SUBJECT'S LAST NAME: IANNONE

DATE/TIME OF TEST: 01/24/2016 21:13:55

Diagnostics Check		OK	21:05
Air Blank		0.000	21:05
Control Test		0.078	21:06
Air Blank		0.000	21:06
Subject Sample	# 1	0.173	21:08
Breath Volume		1.968	21:08
Air Blank		0.000	21:09
Air Blank		0.000	21:11
Subject Sample	# 2	0.166	21:12
Breath Volume		2.500	21:12
Air Blank		0.000	21:12
Control Test		0.076	21:13
Air Blank		0.000	21:13
Diagnostics Check		OK	21:13

Data Download Date/Time: 01/30/2016 15:10

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PESCO,MATTHEW,J
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CLARK,BRENT
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/25/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 20:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:34

SUBJECT'S LAST NAME: PENSO

DATE/TIME OF TEST: 01/25/2016 21:00:59

Diagnostics Check		OK	20:54
Air Blank		0.000	20:55
Control Test		0.078	20:55
Air Blank		0.000	20:55
Subject Sample	# 1	0.236	20:56
Breath Volume		2.484	20:56
Air Blank		0.000	20:56
Air Blank		0.000	20:58
Subject Sample	# 2	0.251	20:59
Breath Volume		2.925	20:59
Air Blank		0.000	21:00
Control Test		0.078	21:00
Air Blank		0.000	21:00
Diagnostics Check		OK	21:00

Data Download Date/Time: 01/30/2016 15:11

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTER,MATTHEW,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ALVEY,DOUGLAS
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/25/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 19:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:54

SUBJECT'S LAST NAME: CHRISTIANSEN

DATE/TIME OF TEST: 01/27/2016 19:24:49

Diagnostics Check		OK	19:16	
Air Blank		0.000	19:16	
Control Test		0.078	19:17	
Air Blank		0.000	19:17	
Subject Sample	# 1		19:20	No Sample Provided
Breath Volume		0.000	19:20	
Air Blank		0.000	19:21	
Air Blank		0.000	19:23	
Subject Sample	# 2		19:23	Subject Test Refused
Breath Volume		0.000	19:23	
Air Blank		0.000	19:23	
Control Test		0.077	19:24	
Air Blank		0.000	19:24	
Diagnostics Check		OK	19:24	

Data Download Date/Time: 01/30/2016 15:11

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALVEY,DOUGLAS,J
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ZOLLOTA,VLADAMIR
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/25/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 20:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:24

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 01/29/2016 20:58:08

Diagnostics Check		OK	20:51
Air Blank		0.000	20:52
Control Test		0.078	20:52
Air Blank		0.000	20:53
Subject Sample	# 1	0.149	20:53
Breath Volume		2.566	20:53
Air Blank		0.000	20:54
Air Blank		0.000	20:56
Subject Sample	# 2	0.147	20:56
Breath Volume		2.214	20:56
Air Blank		0.000	20:57
Control Test		0.079	20:57
Air Blank		0.000	20:58
Diagnostics Check		OK	20:58

Data Download Date/Time: 01/30/2016 15:11

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTER,MATTHEW,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ALVEY,DOUGLAS
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/25/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 01:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:43

SUBJECT'S LAST NAME: DUGAN

DATE/TIME OF TEST: 01/30/2016 01:19:33

Diagnostics Check		OK	01:13
Air Blank		0.000	01:13
Control Test		0.079	01:13
Air Blank		0.000	01:14
Subject Sample	# 1	0.234	01:14
Breath Volume		2.191	01:14
Air Blank		0.000	01:15
Air Blank		0.000	01:17
Subject Sample	# 2	0.237	01:18
Breath Volume		2.378	01:18
Air Blank		0.000	01:18
Control Test		0.078	01:18
Air Blank		0.000	01:19
Diagnostics Check		OK	01:19

Data Download Date/Time: 01/30/2016 15:11

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTER,MATTHEW,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ZOLLOTA,VLADIMIR
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/25/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 01:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: STEELE

DATE/TIME OF TEST: 01/30/2016 01:48:37

Diagnostics Check		OK	01:45	
Air Blank		0.000	01:46	
Control Test		0.077	01:46	
Air Blank		0.000	01:47	
Subject Sample	# 1		01:47	Subject Test Refused
Breath Volume		0.000	01:47	
Air Blank		0.000	01:47	
Control Test		0.079	01:48	
Air Blank		0.000	01:48	
Diagnostics Check		OK	01:48	

Data Download Date/Time: 01/30/2016 15:11

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTER,MATTHEW,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ZOLLOTA,VLADIMIR
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 01/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 20:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:33

SUBJECT'S LAST NAME: DUERST

DATE/TIME OF TEST: 01/30/2016 20:58:35

Diagnostics Check		OK	20:54	
Air Blank		0.000	20:55	
Control Test		0.079	20:55	
Air Blank		0.000	20:56	
Subject Sample	# 1		20:57	Subject Test Refused
Breath Volume		0.000	20:57	
Air Blank		0.000	20:57	
Control Test		0.079	20:57	
Air Blank		0.000	20:58	
Diagnostics Check		OK	20:58	

Data Download Date/Time: 02/24/2016 22:09

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HEALY,G,B
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: HEALY,G
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 12/14/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A3
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:41

SUBJECT'S LAST NAME: SILVA

DATE/TIME OF TEST: 01/01/2016 02:12:54

Diagnostics Check		OK	02:06
Air Blank		0.000	02:07
Control Test		0.080	02:07
Air Blank		0.000	02:08
Subject Sample	# 1	0.159	02:08
Breath Volume		2.757	02:08
Air Blank		0.000	02:09
Air Blank		0.000	02:11
Subject Sample	# 2	0.147	02:11
Breath Volume		2.460	02:11
Air Blank		0.000	02:11
Control Test		0.080	02:12
Air Blank		0.000	02:12
Diagnostics Check		OK	02:12

Data Download Date/Time: 01/27/2016 05:28

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROE,ROBERT,D
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: GARRISON,JOHN
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 12/19/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: BANASIK

DATE/TIME OF TEST: 01/01/2016 00:24:24

Diagnostics Check		OK	00:17
Air Blank		0.000	00:17
Control Test		0.079	00:18
Air Blank		0.000	00:18
Subject Sample	# 1	0.096	00:19
Breath Volume		1.339	00:19
Air Blank		0.000	00:20
Air Blank		0.000	00:22
Subject Sample	# 2	0.096	00:22
Breath Volume		1.500	00:22
Air Blank		0.000	00:23
Control Test		0.080	00:23
Air Blank		0.000	00:24
Diagnostics Check		OK	00:24

Data Download Date/Time: 01/15/2016 14:05

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRACKETT,MARK,O
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: MILLER,PAUL
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 12/19/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: MCCRAY

DATE/TIME OF TEST: 01/03/2016 03:31:16

Diagnostics Check		OK	03:28	
Air Blank		0.000	03:28	
Control Test		0.080	03:28	
Air Blank		0.000	03:29	
Subject Sample	# 1		03:29	Subject Test Refused
Breath Volume		0.000	03:29	
Air Blank		0.000	03:30	
Control Test		0.076	03:30	
Air Blank		0.000	03:31	
Diagnostics Check		OK	03:31	

Data Download Date/Time: 01/15/2016 14:05

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPENCER,SHANE,
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: SOTHERLUND,NICK
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 12/19/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 19:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:17

SUBJECT'S LAST NAME: GONSALVES

DATE/TIME OF TEST: 01/05/2016 20:47:15

Diagnostics Check		OK	20:40
Air Blank		0.000	20:40
Control Test		0.080	20:40
Air Blank		0.000	20:41
Subject Sample	# 1	0.000	20:42
Breath Volume		1.660	20:42
Air Blank		0.000	20:42
Air Blank		0.000	20:44
Subject Sample	# 2	0.000	20:45
Breath Volume		1.625	20:45
Air Blank		0.000	20:46
Control Test		0.079	20:46
Air Blank		0.000	20:47
Diagnostics Check		OK	20:47

Data Download Date/Time: 01/15/2016 14:05

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,SUSAN,L
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: SMITH,SUSAN
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 01/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: COURT ORD
ARREST TIME: 14:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:00

SUBJECT'S LAST NAME: RICHARDSON

DATE/TIME OF TEST: 01/15/2016 16:41:10

Diagnostics Check		OK	16:35
Air Blank		0.000	16:35
Control Test		0.078	16:36
Air Blank		0.000	16:36
Subject Sample	# 1	0.000	16:36
Breath Volume		2.535	16:36
Air Blank		0.000	16:37
Air Blank		0.000	16:39
Subject Sample	# 2	0.000	16:39
Breath Volume		2.609	16:39
Air Blank		0.000	16:40
Control Test		0.077	16:40
Air Blank		0.000	16:41
Diagnostics Check		OK	16:41

Data Download Date/Time: 02/19/2016 13:20

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,SUSAN,L
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: SMITH,SUSAN
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 01/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: COURT ORD
ARREST TIME: 16:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:30

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 01/15/2016 17:02:18

Diagnostics Check		OK	16:56
Air Blank		0.000	16:56
Control Test		0.079	16:57
Air Blank		0.000	16:57
Subject Sample	# 1	0.000	16:58
Breath Volume		1.878	16:58
Air Blank		0.000	16:58
Air Blank		0.000	17:00
Subject Sample	# 2	0.000	17:00
Breath Volume		1.363	17:00
Air Blank		0.000	17:01
Control Test		0.078	17:01
Air Blank		0.000	17:02
Diagnostics Check		OK	17:02

Data Download Date/Time: 02/19/2016 13:20

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,SUSAN,L
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: SMITH,SUSAN
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 01/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: COURT ORD
ARREST TIME: 16:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:30

SUBJECT'S LAST NAME: NOLAND

DATE/TIME OF TEST: 01/15/2016 17:12:10

Diagnostics Check		OK	17:06
Air Blank		0.000	17:06
Control Test		0.079	17:07
Air Blank		0.000	17:07
Subject Sample	# 1	0.000	17:08
Breath Volume		2.718	17:08
Air Blank		0.000	17:08
Air Blank		0.000	17:10
Subject Sample	# 2	0.000	17:10
Breath Volume		2.738	17:10
Air Blank		0.000	17:11
Control Test		0.080	17:11
Air Blank		0.000	17:12
Diagnostics Check		OK	17:12

Data Download Date/Time: 02/19/2016 13:20

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,SUSAN,L
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: SMITH,SUSAN
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 01/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: COURT ORD
ARREST TIME: 16:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:45

SUBJECT'S LAST NAME: NOLAND

DATE/TIME OF TEST: 01/15/2016 17:21:49

Diagnostics Check		OK	17:15
Air Blank		0.000	17:16
Control Test		0.079	17:16
Air Blank		0.000	17:17
Subject Sample	# 1	0.000	17:17
Breath Volume		2.984	17:17
Air Blank		0.000	17:18
Air Blank		0.000	17:20
Subject Sample	# 2	0.000	17:20
Breath Volume		2.878	17:20
Air Blank		0.000	17:20
Control Test		0.078	17:21
Air Blank		0.000	17:21
Diagnostics Check		OK	17:21

Data Download Date/Time: 02/19/2016 13:20

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORSE,MICHAEL,P
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: PTR,PTR
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 01/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: COURT ORD
ARREST TIME: 09:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:03

SUBJECT'S LAST NAME: SALDIVAR-SLAUGHTER

DATE/TIME OF TEST: 01/16/2016 09:28:37

Diagnostics Check		OK	09:25	
Air Blank		0.000	09:25	
Control Test		0.079	09:25	
Air Blank		0.000	09:26	
Subject Sample	# 1	0.000	09:26	
Breath Volume		2.804	09:26	
Air Blank			09:27	RFI Detect
Air Blank		0.000	09:27	
Control Test		0.079	09:28	
Air Blank		0.000	09:28	
Diagnostics Check		OK	09:28	

Data Download Date/Time: 02/19/2016 13:20

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORSE,MICHAEL,P
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: PTR,PTR
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 01/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: COURT ORD
ARREST TIME: 09:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:00

SUBJECT'S LAST NAME: SALDIVAR-SLAUGHTER

DATE/TIME OF TEST: 01/16/2016 09:39:05

Diagnostics Check		OK	09:32
Air Blank		0.000	09:33
Control Test		0.078	09:33
Air Blank		0.000	09:34
Subject Sample	# 1	0.000	09:34
Breath Volume		2.617	09:34
Air Blank		0.000	09:35
Air Blank		0.000	09:37
Subject Sample	# 2	0.000	09:37
Breath Volume		3.050	09:37
Air Blank		0.000	09:38
Control Test		0.079	09:38
Air Blank		0.000	09:38
Diagnostics Check		OK	09:39

Data Download Date/Time: 02/19/2016 13:21

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORSE,MICHAEL,P
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: RITENOUR,STEVEN
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 01/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 06:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:47

SUBJECT'S LAST NAME: HANKERSON

DATE/TIME OF TEST: 01/17/2016 07:14:52

Diagnostics Check		OK	07:08
Air Blank		0.000	07:08
Control Test		0.078	07:09
Air Blank		0.000	07:09
Subject Sample	# 1	0.189	07:10
Breath Volume		2.417	07:10
Air Blank		0.000	07:10
Air Blank		0.000	07:12
Subject Sample	# 2	0.178	07:13
Breath Volume		2.113	07:13
Air Blank		0.000	07:13
Control Test		0.078	07:14
Air Blank		0.000	07:14
Diagnostics Check		OK	07:14

Data Download Date/Time: 02/19/2016 13:21

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ABELN,SHON,W
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: MILLER,PAUL
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 01/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:07

SUBJECT'S LAST NAME: HICKS

DATE/TIME OF TEST: 01/22/2016 04:34:17

Diagnostics Check		OK	04:28
Air Blank		0.000	04:28
Control Test		0.079	04:28
Air Blank		0.000	04:29
Subject Sample	# 1	0.104	04:29
Breath Volume		2.078	04:29
Air Blank		0.000	04:30
Air Blank		0.000	04:32
Subject Sample	# 2	0.112	04:32
Breath Volume		2.140	04:32
Air Blank		0.000	04:33
Control Test		0.080	04:33
Air Blank		0.000	04:34
Diagnostics Check		OK	04:34

Data Download Date/Time: 02/19/2016 13:21

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUNN,T,M
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: DOUBERLEY,CHAD
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 01/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:30

SUBJECT'S LAST NAME: ARMSTRONG

DATE/TIME OF TEST: 01/22/2016 05:57:51

Diagnostics Check		OK	05:52	
Air Blank		0.000	05:52	
Control Test		0.079	05:53	
Air Blank		0.000	05:53	
Subject Sample	# 1		05:56	Subject Test Refused
Breath Volume		0.000	05:56	
Air Blank		0.000	05:56	
Control Test		0.079	05:57	
Air Blank		0.000	05:57	
Diagnostics Check		OK	05:57	

Data Download Date/Time: 02/19/2016 13:21

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MENDEZ,ROBERT,C
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: MENDEZ,ROBERT
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/02/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07014080A2
DRY GAS EXPIRY DATE: 05/01/2016
VIOLATION CODE: DUI
ARREST TIME: 23:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:59

SUBJECT'S LAST NAME: THURSTON

DATE/TIME OF TEST: 01/01/2016 00:31:11

Diagnostics Check		OK	00:21	
Air Blank		0.000	00:21	
Control Test		0.077	00:22	
Air Blank		0.000	00:22	
Subject Sample	# 1		00:25	No Sample Provided
Breath Volume		0.000	00:25	
Air Blank		0.000	00:26	
Air Blank		0.000	00:28	
Subject Sample	# 2		00:29	Subject Test Refused
Breath Volume		0.000	00:29	
Air Blank		0.000	00:30	
Control Test		0.077	00:30	
Air Blank		0.000	00:31	
Diagnostics Check		OK	00:31	

Data Download Date/Time: 01/11/2016 13:25

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001056

INSTRUMENT REGISTERED TO: DAVIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAINES,DAVID,K
BREATH TEST OPERATOR AGENCY: Davie PD
ARRESTING OFFICER: YOUNG,JOSHUA
ARRESTING OFFICER AGENCY: Davie PD
LAST AGENCY INSPECTION: 01/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG429701
DRY GAS EXPIRY DATE: 10/24/2016
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: MERCADO

DATE/TIME OF TEST: 01/17/2016 02:53:43

Diagnostics Check		OK	02:46
Air Blank		0.000	02:47
Control Test		0.079	02:47
Air Blank		0.000	02:48
Subject Sample	# 1	0.129	02:48
Breath Volume		1.253	02:48
Air Blank		0.000	02:49
Air Blank		0.000	02:51
Subject Sample	# 2	0.122	02:52
Breath Volume		1.167	02:52
Air Blank		0.000	02:52
Control Test		0.079	02:53
Air Blank		0.000	02:53
Diagnostics Check		OK	02:53

Data Download Date/Time: 03/04/2016 10:58

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001056

INSTRUMENT REGISTERED TO: DAVIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,DANIEL,
BREATH TEST OPERATOR AGENCY: Davie PD
ARRESTING OFFICER: CAVALIE,YANNIS
ARRESTING OFFICER AGENCY: Davie PD
LAST AGENCY INSPECTION: 01/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG429701
DRY GAS EXPIRY DATE: 10/24/2016
VIOLATION CODE: DUI
ARREST TIME: 20:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: TOMMIE

DATE/TIME OF TEST: 01/27/2016 21:49:17

Diagnostics Check		OK	21:42
Air Blank		0.000	21:43
Control Test		0.080	21:43
Air Blank		0.000	21:44
Subject Sample	# 1	0.182	21:44
Breath Volume		2.804	21:44
Air Blank		0.000	21:45
Air Blank		0.000	21:47
Subject Sample	# 2	0.179	21:47
Breath Volume		2.753	21:47
Air Blank		0.000	21:48
Control Test		0.079	21:48
Air Blank		0.000	21:49
Diagnostics Check		OK	21:49

Data Download Date/Time: 03/04/2016 10:58

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001056

INSTRUMENT REGISTERED TO: DAVIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAINES,DAVID,K
BREATH TEST OPERATOR AGENCY: Davie PD
ARRESTING OFFICER: HAINES,DAVID
ARRESTING OFFICER AGENCY: Davie PD
LAST AGENCY INSPECTION: 01/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG429701
DRY GAS EXPIRY DATE: 10/24/2016
VIOLATION CODE: DUI
ARREST TIME: 23:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: OLSON

DATE/TIME OF TEST: 01/31/2016 00:32:46

Diagnostics Check		OK	00:26
Air Blank		0.000	00:26
Control Test		0.080	00:26
Air Blank		0.000	00:27
Subject Sample	# 1	0.243	00:28
Breath Volume		3.433	00:28
Air Blank		0.000	00:28
Air Blank		0.000	00:30
Subject Sample	# 2	0.252	00:31
Breath Volume		3.484	00:31
Air Blank		0.000	00:31
Control Test		0.079	00:32
Air Blank		0.000	00:32
Diagnostics Check		OK	00:32

Data Download Date/Time: 03/04/2016 10:58

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001056

INSTRUMENT REGISTERED TO: DAVIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAINES,DAVID,K
BREATH TEST OPERATOR AGENCY: Davie PD
ARRESTING OFFICER: YOUNG,JOSHUA
ARRESTING OFFICER AGENCY: Davie PD
LAST AGENCY INSPECTION: 01/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG429701
DRY GAS EXPIRY DATE: 10/24/2016
VIOLATION CODE: DUI
ARREST TIME: 04:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:43

SUBJECT'S LAST NAME: COVER

DATE/TIME OF TEST: 01/31/2016 05:09:37

Diagnostics Check		OK	05:03
Air Blank		0.000	05:03
Control Test		0.080	05:03
Air Blank		0.000	05:04
Subject Sample	# 1	0.149	05:05
Breath Volume		1.390	05:05
Air Blank		0.000	05:05
Air Blank		0.000	05:07
Subject Sample	# 2	0.157	05:08
Breath Volume		2.023	05:08
Air Blank		0.000	05:08
Control Test		0.080	05:09
Air Blank		0.000	05:09
Diagnostics Check		OK	05:09

Data Download Date/Time: 03/04/2016 10:58

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001057

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KELLNER,MICHAEL,A
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: LANDRY,CHAD
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 12/15/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:46

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 01/17/2016 06:44:45

Diagnostics Check		OK	06:36
Air Blank		0.000	06:37
Control Test		0.078	06:37
Air Blank		0.000	06:37
Subject Sample	# 1	0.202	06:40
Breath Volume		1.769	06:40
Air Blank		0.000	06:40
Air Blank		0.000	06:42
Subject Sample	# 2	0.208	06:43
Breath Volume		1.261	06:43
Air Blank		0.000	06:43
Control Test		0.077	06:44
Air Blank		0.000	06:44
Diagnostics Check		OK	06:44

Data Download Date/Time: 01/25/2016 07:45

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001057

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUACKENBUSH,BRET,C
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: QUACKENBUSH,BRET
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 12/15/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:34

SUBJECT'S LAST NAME: STEPP

DATE/TIME OF TEST: 01/18/2016 04:04:26

Diagnostics Check		OK	03:55	
Air Blank		0.000	03:56	
Control Test		0.080	03:56	
Air Blank		0.000	03:57	
Subject Sample	# 1		04:00	No Sample Provided
Breath Volume		0.000	04:00	
Air Blank		0.000	04:00	
Air Blank		0.000	04:02	
Subject Sample	# 2		04:03	Subject Test Refused
Breath Volume		0.000	04:03	
Air Blank		0.000	04:03	
Control Test		0.080	04:03	
Air Blank		0.000	04:04	
Diagnostics Check		OK	04:04	

Data Download Date/Time: 01/25/2016 07:45

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001057

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTLE,CORY,A
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: SUTTLE,CORY
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 12/15/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:13

SUBJECT'S LAST NAME: KLUE

DATE/TIME OF TEST: 01/24/2016 22:41:03

Diagnostics Check		OK	22:33
Air Blank		0.000	22:34
Control Test		0.080	22:34
Air Blank		0.000	22:35
Subject Sample	# 1	0.196	22:36
Breath Volume		2.625	22:36
Air Blank		0.000	22:37
Air Blank		0.000	22:38
Subject Sample	# 2	0.190	22:39
Breath Volume		2.234	22:39
Air Blank		0.000	22:40
Control Test		0.079	22:40
Air Blank		0.000	22:40
Diagnostics Check		OK	22:41

Data Download Date/Time: 01/25/2016 07:46

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001059

INSTRUMENT REGISTERED TO: FERNANDINA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAZURYK,MICHAEL,S
BREATH TEST OPERATOR AGENCY: Fernandina Beach PD
ARRESTING OFFICER: MAZURYK,MICHAEL
ARRESTING OFFICER AGENCY: Fernandina Beach PD
LAST AGENCY INSPECTION: 12/09/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A3
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:14

SUBJECT'S LAST NAME: THORNTON

DATE/TIME OF TEST: 01/01/2016 04:47:08

Diagnostics Check		OK	04:40
Air Blank		0.000	04:41
Control Test		0.079	04:41
Air Blank		0.000	04:42
Subject Sample	# 1	0.000	04:42
Breath Volume		4.152	04:42
Air Blank		0.000	04:43
Air Blank		0.000	04:45
Subject Sample	# 2	0.000	04:45
Breath Volume		2.449	04:45
Air Blank		0.000	04:46
Control Test		0.079	04:46
Air Blank		0.000	04:47
Diagnostics Check		OK	04:47

Data Download Date/Time: 01/06/2016 11:41

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001059

INSTRUMENT REGISTERED TO: FERNANDINA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,JORGE,L
BREATH TEST OPERATOR AGENCY: Fernandina Beach PD
ARRESTING OFFICER: PAETSCH,JOSHUA
ARRESTING OFFICER AGENCY: Fernandina Beach PD
LAST AGENCY INSPECTION: 01/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A3
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:17

SUBJECT'S LAST NAME: TYO

DATE/TIME OF TEST: 01/30/2016 02:46:38

Diagnostics Check		OK	02:40	
Air Blank		0.000	02:41	
Control Test		0.080	02:41	
Air Blank		0.000	02:41	
Subject Sample	# 1	0.174	02:45	Volume Not Met
Breath Volume		0.285	02:45	
Air Blank			02:45	Purge Fail
Air Blank			02:46	Purge Fail

Data Download Date/Time: 02/16/2016 11:03

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001059

INSTRUMENT REGISTERED TO: FERNANDINA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,JORGE,L
BREATH TEST OPERATOR AGENCY: Fernandina Beach PD
ARRESTING OFFICER: PAETSCH,JOSHUA
ARRESTING OFFICER AGENCY: Fernandina Beach PD
LAST AGENCY INSPECTION: 01/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A3
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:17

SUBJECT'S LAST NAME: TYO

DATE/TIME OF TEST: 01/30/2016 03:02:59

Diagnostics Check		OK	02:49	
Air Blank		0.000	02:49	
Control Test		0.079	02:49	
Air Blank		0.000	02:50	
Subject Sample	# 1	0.184	02:53	Volume Not Met
Breath Volume		0.015	02:53	
Air Blank		0.000	02:54	
Air Blank		0.000	02:55	
Subject Sample	# 2	0.164	02:56	
Breath Volume		1.242	02:56	
Air Blank		0.000	02:57	
Air Blank		0.000	02:58	
Subject Sample	# 3	0.173	03:01	
Breath Volume		1.378	03:01	
Air Blank		0.000	03:02	
Control Test		0.079	03:02	
Air Blank		0.000	03:02	
Diagnostics Check		OK	03:02	

Data Download Date/Time: 02/16/2016 11:03

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001060

INSTRUMENT REGISTERED TO: WINTER HAVEN PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHERMAN,MATTHEW,P
BREATH TEST OPERATOR AGENCY: Winter Haven PD
ARRESTING OFFICER: SHERMAN,MATTHEW
ARRESTING OFFICER AGENCY: Winter Haven PD
LAST AGENCY INSPECTION: 12/31/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14214080A6
DRY GAS EXPIRY DATE: 07/05/2016
VIOLATION CODE: DUI
ARREST TIME: 02:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:32

SUBJECT'S LAST NAME: NELSON

DATE/TIME OF TEST: 01/16/2016 04:06:37

Diagnostics Check		OK	03:59
Air Blank		0.000	04:00
Control Test		0.080	04:00
Air Blank		0.000	04:01
Subject Sample	# 1	0.084	04:01
Breath Volume		2.332	04:01
Air Blank		0.000	04:02
Air Blank		0.000	04:04
Subject Sample	# 2	0.083	04:04
Breath Volume		2.519	04:04
Air Blank		0.000	04:05
Control Test		0.080	04:05
Air Blank		0.000	04:06
Diagnostics Check		OK	04:06

Data Download Date/Time: 01/28/2016 12:26

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001060

INSTRUMENT REGISTERED TO: WINTER HAVEN PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NGUYEN,JOHN,
BREATH TEST OPERATOR AGENCY: Winter Haven PD
ARRESTING OFFICER: NGUYEN,JOHN
ARRESTING OFFICER AGENCY: Winter Haven PD
LAST AGENCY INSPECTION: 12/31/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14214080A6
DRY GAS EXPIRY DATE: 07/05/2016
VIOLATION CODE: DUI
ARREST TIME: 21:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:34

SUBJECT'S LAST NAME: MAYA

DATE/TIME OF TEST: 01/26/2016 22:06:16

Diagnostics Check		OK	21:57	
Air Blank		0.000	21:57	
Control Test		0.080	21:58	
Air Blank		0.000	21:58	
Subject Sample	# 1		22:01	No Sample Provided
Breath Volume		0.000	22:01	
Air Blank		0.000	22:02	
Air Blank		0.000	22:04	
Subject Sample	# 2		22:04	Subject Test Refused
Breath Volume		0.000	22:04	
Air Blank		0.000	22:05	
Control Test		0.080	22:05	
Air Blank		0.000	22:06	
Diagnostics Check		OK	22:06	

Data Download Date/Time: 01/28/2016 12:26

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001060

INSTRUMENT REGISTERED TO: WINTER HAVEN PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NGUYEN,JOHN,
BREATH TEST OPERATOR AGENCY: Winter Haven PD
ARRESTING OFFICER: NGUYEN,JOHN
ARRESTING OFFICER AGENCY: Winter Haven PD
LAST AGENCY INSPECTION: 01/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14214080A6
DRY GAS EXPIRY DATE: 07/05/2016
VIOLATION CODE: DUI
ARREST TIME: 02:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:19

SUBJECT'S LAST NAME: DORITY

DATE/TIME OF TEST: 01/31/2016 02:43:56

Diagnostics Check		OK	02:40	
Air Blank		0.000	02:41	
Control Test		0.081	02:41	
Air Blank		0.000	02:42	
Subject Sample	# 1		02:42	Subject Test Refused
Breath Volume		0.000	02:42	
Air Blank		0.000	02:42	
Control Test		0.080	02:43	
Air Blank		0.000	02:43	
Diagnostics Check		OK	02:43	

Data Download Date/Time: 02/26/2016 00:29

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001062

INSTRUMENT REGISTERED TO: TREASURE ISLAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CONTOIS,R,
BREATH TEST OPERATOR AGENCY: Treasure Island PD
ARRESTING OFFICER: CONTOIS,R
ARRESTING OFFICER AGENCY: Treasure Island PD
LAST AGENCY INSPECTION: 12/16/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20314080A5
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 00:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:51

SUBJECT'S LAST NAME: GIBSON JR

DATE/TIME OF TEST: 01/01/2016 01:37:30

Diagnostics Check		OK	01:29
Air Blank		0.000	01:30
Control Test		0.079	01:30
Air Blank		0.000	01:30
Subject Sample	# 1	0.184	01:32
Breath Volume		2.761	01:32
Air Blank		0.000	01:33
Air Blank		0.000	01:35
Subject Sample	# 2	0.177	01:35
Breath Volume		2.597	01:35
Air Blank		0.000	01:36
Control Test		0.079	01:36
Air Blank		0.000	01:37
Diagnostics Check		OK	01:37

Data Download Date/Time: 01/19/2016 09:03

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001062

INSTRUMENT REGISTERED TO: TREASURE ISLAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CONTOIS,R,
BREATH TEST OPERATOR AGENCY: Treasure Island PD
ARRESTING OFFICER: SMITH,M
ARRESTING OFFICER AGENCY: Treasure Island PD
LAST AGENCY INSPECTION: 01/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20314080A5
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 01:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:51

SUBJECT'S LAST NAME: KINNEY

DATE/TIME OF TEST: 01/17/2016 02:23:47

Diagnostics Check		OK	02:17
Air Blank		0.000	02:17
Control Test		0.080	02:18
Air Blank		0.000	02:18
Subject Sample	# 1	0.192	02:19
Breath Volume		1.539	02:19
Air Blank		0.000	02:19
Air Blank		0.000	02:21
Subject Sample	# 2	0.190	02:22
Breath Volume		1.640	02:22
Air Blank		0.000	02:22
Control Test		0.080	02:23
Air Blank		0.000	02:23
Diagnostics Check		OK	02:23

Data Download Date/Time: 01/19/2016 09:03

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001063

INSTRUMENT REGISTERED TO: HOLLYWOOD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STUDENMUND,ERICA,C
BREATH TEST OPERATOR AGENCY: Hollywood PD
ARRESTING OFFICER: HAZEL,MARK
ARRESTING OFFICER AGENCY: Hollywood PD
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 01:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: VAUGHN

DATE/TIME OF TEST: 01/01/2016 02:50:26

Diagnostics Check		OK	02:44
Air Blank		0.000	02:44
Control Test		0.080	02:45
Air Blank		0.000	02:45
Subject Sample	# 1	0.077	02:46
Breath Volume		1.460	02:46
Air Blank		0.000	02:46
Air Blank		0.000	02:48
Subject Sample	# 2	0.072	02:48
Breath Volume		1.277	02:48
Air Blank		0.000	02:49
Control Test		0.079	02:49
Air Blank		0.000	02:50
Diagnostics Check		OK	02:50

Data Download Date/Time: 01/29/2016 18:35

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001063

INSTRUMENT REGISTERED TO: HOLLYWOOD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DERIET,RAFAEL,
BREATH TEST OPERATOR AGENCY: Hollywood PD
ARRESTING OFFICER: DERIET,RAFAEL
ARRESTING OFFICER AGENCY: Hollywood PD
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 23:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: LAWSON

DATE/TIME OF TEST: 01/04/2016 00:28:09

Diagnostics Check		OK	00:21
Air Blank		0.000	00:22
Control Test		0.080	00:22
Air Blank		0.000	00:22
Subject Sample	# 1	0.189	00:23
Breath Volume		1.148	00:23
Air Blank		0.000	00:23
Air Blank		0.000	00:25
Subject Sample	# 2	0.190	00:26
Breath Volume		1.640	00:26
Air Blank		0.000	00:27
Control Test		0.081	00:27
Air Blank		0.000	00:28
Diagnostics Check		OK	00:28

Data Download Date/Time: 01/29/2016 18:35

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001063

INSTRUMENT REGISTERED TO: HOLLYWOOD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEO,JUSTIN,B
BREATH TEST OPERATOR AGENCY: Hollywood PD
ARRESTING OFFICER: HERRERA,M
ARRESTING OFFICER AGENCY: Hollywood PD
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 18:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:27

SUBJECT'S LAST NAME: KHRINO

DATE/TIME OF TEST: 01/10/2016 19:50:12

Diagnostics Check		OK	19:47	
Air Blank		0.000	19:47	
Control Test		0.080	19:48	
Air Blank		0.000	19:48	
Subject Sample	# 1		19:48	Subject Test Refused
Breath Volume		0.000	19:48	
Air Blank		0.000	19:49	
Control Test		0.080	19:49	
Air Blank		0.000	19:50	
Diagnostics Check		OK	19:50	

Data Download Date/Time: 01/29/2016 18:36

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001063

INSTRUMENT REGISTERED TO: HOLLYWOOD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DERIET,RAFAEL,
BREATH TEST OPERATOR AGENCY: Hollywood PD
ARRESTING OFFICER: BALUJA,JONATHAN
ARRESTING OFFICER AGENCY: Hollywood PD
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 04:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:24

SUBJECT'S LAST NAME: MCARTHUR

DATE/TIME OF TEST: 01/21/2016 05:00:28

Diagnostics Check		OK	04:44	
Air Blank		0.000	04:44	
Control Test		0.080	04:45	
Air Blank		0.000	04:45	
Subject Sample	# 1	0.209	04:48	Volume Not Met
Breath Volume		0.316	04:48	
Air Blank		0.000	04:49	
Air Blank		0.000	04:51	
Subject Sample	# 2	0.242	04:54	
Breath Volume		1.707	04:54	
Air Blank		0.000	04:55	
Air Blank		0.000	04:56	
Subject Sample	# 3	0.237	04:58	
Breath Volume		1.125	04:58	
Air Blank		0.000	04:59	
Control Test		0.079	04:59	
Air Blank		0.000	05:00	
Diagnostics Check		OK	05:00	

Data Download Date/Time: 01/29/2016 18:36

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001063

INSTRUMENT REGISTERED TO: HOLLYWOOD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DERIET,RAFAEL,
BREATH TEST OPERATOR AGENCY: Hollywood PD
ARRESTING OFFICER: DERIET,RAFAEL
ARRESTING OFFICER AGENCY: Hollywood PD
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 23:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: RUIZ-COLINDRES

DATE/TIME OF TEST: 01/23/2016 00:26:49

Diagnostics Check	OK	00:25	
Air Blank	0.000	00:26	
Control Test	0.000	00:26	RFI Detect
Air Blank	0.000	00:26	

Data Download Date/Time: 01/29/2016 18:36

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001063

INSTRUMENT REGISTERED TO: HOLLYWOOD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DERIET,RAFAEL,
BREATH TEST OPERATOR AGENCY: Hollywood PD
ARRESTING OFFICER: DERIET,RAFAEL
ARRESTING OFFICER AGENCY: Hollywood PD
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 23:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: RUIZ-COLINDRES

DATE/TIME OF TEST: 01/23/2016 00:32:47

Diagnostics Check		OK	00:29	
Air Blank		0.000	00:30	
Control Test		0.081	00:30	
Air Blank		0.000	00:31	
Subject Sample	# 1		00:31	RFI Detect
Breath Volume		0.000	00:31	
Air Blank		0.000	00:31	
Control Test		0.081	00:32	
Air Blank		0.000	00:32	
Diagnostics Check		OK	00:32	

Data Download Date/Time: 01/29/2016 18:36

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001063

INSTRUMENT REGISTERED TO: HOLLYWOOD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DERIET,RAFAEL,
BREATH TEST OPERATOR AGENCY: Hollywood PD
ARRESTING OFFICER: DERIET,RAFAEL
ARRESTING OFFICER AGENCY: Hollywood PD
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 23:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: RUIZ-COLINDRES

DATE/TIME OF TEST: 01/23/2016 00:41:36

Diagnostics Check		OK	00:34
Air Blank		0.000	00:35
Control Test		0.081	00:35
Air Blank		0.000	00:36
Subject Sample	# 1	0.206	00:37
Breath Volume		1.996	00:37
Air Blank		0.000	00:37
Air Blank		0.000	00:39
Subject Sample	# 2	0.195	00:40
Breath Volume		1.875	00:40
Air Blank		0.000	00:40
Control Test		0.081	00:41
Air Blank		0.000	00:41
Diagnostics Check		OK	00:41

Data Download Date/Time: 01/29/2016 18:36

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001063

INSTRUMENT REGISTERED TO: HOLLYWOOD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DERIET,RAFAEL,
BREATH TEST OPERATOR AGENCY: Hollywood PD
ARRESTING OFFICER: DERIET,RAFAEL
ARRESTING OFFICER AGENCY: Hollywood PD
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 22:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:09

SUBJECT'S LAST NAME: RAMOS

DATE/TIME OF TEST: 01/23/2016 01:00:38

Diagnostics Check		OK	00:50
Air Blank		0.000	00:51
Control Test		0.081	00:51
Air Blank		0.000	00:52
Subject Sample	# 1	0.194	00:53
Breath Volume		1.523	00:53
Air Blank		0.000	00:54
Air Blank		0.000	00:56
Subject Sample	# 2	0.194	00:59
Breath Volume		1.273	00:59
Air Blank		0.000	00:59
Control Test		0.081	01:00
Air Blank		0.000	01:00
Diagnostics Check		OK	01:00

Data Download Date/Time: 01/29/2016 18:36

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001063

INSTRUMENT REGISTERED TO: HOLLYWOOD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DERIET,RAFAEL,
BREATH TEST OPERATOR AGENCY: Hollywood PD
ARRESTING OFFICER: MARREN,ROBERT
ARRESTING OFFICER AGENCY: Hollywood PD
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:41

SUBJECT'S LAST NAME: MOSIER

DATE/TIME OF TEST: 01/24/2016 01:16:45

Diagnostics Check		OK	01:10
Air Blank		0.000	01:10
Control Test		0.080	01:10
Air Blank		0.000	01:11
Subject Sample	# 1	0.206	01:11
Breath Volume		1.628	01:11
Air Blank		0.000	01:12
Air Blank		0.000	01:14
Subject Sample	# 2	0.207	01:15
Breath Volume		1.894	01:15
Air Blank		0.000	01:15
Control Test		0.080	01:16
Air Blank		0.000	01:16
Diagnostics Check		OK	01:16

Data Download Date/Time: 01/29/2016 18:36

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001063

INSTRUMENT REGISTERED TO: HOLLYWOOD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DERIET,RAFAEL,
BREATH TEST OPERATOR AGENCY: Hollywood PD
ARRESTING OFFICER: DERIET,RAFAEL
ARRESTING OFFICER AGENCY: Hollywood PD
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 02:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:58

SUBJECT'S LAST NAME: ARRIETA

DATE/TIME OF TEST: 01/24/2016 03:31:01

Diagnostics Check		OK	03:20	
Air Blank		0.000	03:20	
Control Test		0.081	03:20	
Air Blank		0.000	03:21	
Subject Sample	# 1		03:24	No Sample Provided
Breath Volume		0.000	03:24	
Air Blank		0.000	03:25	
Air Blank		0.000	03:27	
Subject Sample	# 2		03:29	Subject Test Refused
Breath Volume		0.000	03:29	
Air Blank		0.000	03:30	
Control Test		0.081	03:30	
Air Blank		0.000	03:30	
Diagnostics Check		OK	03:31	

Data Download Date/Time: 01/29/2016 18:36

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001063

INSTRUMENT REGISTERED TO: HOLLYWOOD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DERIET,RAFAEL,
BREATH TEST OPERATOR AGENCY: Hollywood PD
ARRESTING OFFICER: MEARS,DERRICK
ARRESTING OFFICER AGENCY: Hollywood PD
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 23:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:14

SUBJECT'S LAST NAME: WEST

DATE/TIME OF TEST: 01/29/2016 00:45:31

Diagnostics Check		OK	00:38
Air Blank		0.000	00:39
Control Test		0.081	00:39
Air Blank		0.000	00:40
Subject Sample	# 1	0.179	00:40
Breath Volume		1.445	00:40
Air Blank		0.000	00:41
Air Blank		0.000	00:43
Subject Sample	# 2	0.160	00:44
Breath Volume		3.097	00:44
Air Blank		0.000	00:44
Control Test		0.075	00:44
Air Blank		0.000	00:45
Diagnostics Check		OK	00:45

Data Download Date/Time: 01/29/2016 18:36

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001063

INSTRUMENT REGISTERED TO: HOLLYWOOD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DERIET,RAFAEL,
BREATH TEST OPERATOR AGENCY: Hollywood PD
ARRESTING OFFICER: DERIET,RAFAEL
ARRESTING OFFICER AGENCY: Hollywood PD
LAST AGENCY INSPECTION: 01/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 03:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: GREGORY

DATE/TIME OF TEST: 01/30/2016 03:57:40

Diagnostics Check		OK	03:51
Air Blank		0.000	03:51
Control Test		0.081	03:51
Air Blank		0.000	03:52
Subject Sample	# 1	0.093	03:53
Breath Volume		1.304	03:53
Air Blank		0.000	03:53
Air Blank		0.000	03:55
Subject Sample	# 2	0.094	03:56
Breath Volume		1.355	03:56
Air Blank		0.000	03:56
Control Test		0.081	03:57
Air Blank		0.000	03:57
Diagnostics Check		OK	03:57

Data Download Date/Time: 02/25/2016 17:36

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

INSTRUMENT REGISTERED TO: NEW PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ENGEL,JASON,S
BREATH TEST OPERATOR AGENCY: New Port Richey PD
ARRESTING OFFICER: MCCONNELL,JEFF
ARRESTING OFFICER AGENCY: New Port Richey PD
LAST AGENCY INSPECTION: 01/01/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29614080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 02:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:16

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 01/04/2016 02:45:54

Diagnostics Check		OK	02:39
Air Blank		0.000	02:39
Control Test		0.079	02:40
Air Blank		0.000	02:40
Subject Sample	# 1	0.000	02:41
Breath Volume		1.687	02:41
Air Blank		0.000	02:41
Air Blank		0.000	02:43
Subject Sample	# 2	0.000	02:44
Breath Volume		2.707	02:44
Air Blank		0.000	02:44
Control Test		0.079	02:45
Air Blank		0.000	02:45
Diagnostics Check		OK	02:45

Data Download Date/Time: 02/03/2016 21:33

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

INSTRUMENT REGISTERED TO: NEW PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ENGEL,JASON,S
BREATH TEST OPERATOR AGENCY: New Port Richey PD
ARRESTING OFFICER: EXLINE,ANDREW
ARRESTING OFFICER AGENCY: New Port Richey PD
LAST AGENCY INSPECTION: 01/01/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29614080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 19:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:54

SUBJECT'S LAST NAME: JOERRES

DATE/TIME OF TEST: 01/20/2016 20:24:35

Diagnostics Check		OK	20:17
Air Blank		0.000	20:17
Control Test		0.078	20:18
Air Blank		0.000	20:18
Subject Sample	# 1	0.178	20:19
Breath Volume		1.765	20:19
Air Blank		0.000	20:20
Air Blank		0.000	20:22
Subject Sample	# 2	0.186	20:23
Breath Volume		2.273	20:23
Air Blank		0.000	20:23
Control Test		0.078	20:23
Air Blank		0.000	20:24
Diagnostics Check		OK	20:24

Data Download Date/Time: 02/03/2016 21:33

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

INSTRUMENT REGISTERED TO: NEW PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KELLY,WESLEY,R
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: EXLINE,ANDREW
ARRESTING OFFICER AGENCY: New Port Richey PD
LAST AGENCY INSPECTION: 01/01/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29614080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 01:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: HANSON

DATE/TIME OF TEST: 01/28/2016 02:15:27

Diagnostics Check		OK	02:06
Air Blank		0.000	02:06
Control Test		0.078	02:06
Air Blank		0.000	02:07
Subject Sample	# 1	0.186	02:08
Breath Volume		1.453	02:08
Air Blank		0.000	02:09
Air Blank		0.000	02:11
Subject Sample	# 2	0.186	02:13
Breath Volume		1.687	02:13
Air Blank		0.000	02:14
Control Test		0.078	02:14
Air Blank		0.000	02:15
Diagnostics Check		OK	02:15

Data Download Date/Time: 02/03/2016 21:33

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

INSTRUMENT REGISTERED TO: NEW PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ECK,AARON,R
BREATH TEST OPERATOR AGENCY: New Port Richey PD
ARRESTING OFFICER: EXLINE,ANDREW
ARRESTING OFFICER AGENCY: New Port Richey PD
LAST AGENCY INSPECTION: 01/01/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29614080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 22:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: CIUCCI

DATE/TIME OF TEST: 01/30/2016 23:45:29

Diagnostics Check		OK	23:36	
Air Blank		0.000	23:36	
Control Test		0.078	23:36	
Air Blank		0.000	23:37	
Subject Sample	# 1	0.028	23:40	Volume Not Met
Breath Volume		0.492	23:40	
Air Blank		0.000	23:40	
Air Blank		0.000	23:42	
Subject Sample	# 2		23:44	Subject Test Refused
Breath Volume		0.664	23:44	
Air Blank		0.000	23:44	
Control Test		0.077	23:44	
Air Blank		0.000	23:45	
Diagnostics Check		OK	23:45	

Data Download Date/Time: 02/03/2016 21:33

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001065

INSTRUMENT REGISTERED TO: TARPON SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GASSEN,STEVE,
BREATH TEST OPERATOR AGENCY: Tarpon Springs PD
ARRESTING OFFICER: GASSEN,S
ARRESTING OFFICER AGENCY: Tarpon Springs PD
LAST AGENCY INSPECTION: 12/18/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A2
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 13:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:20

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 01/23/2016 14:47:55

Diagnostics Check		OK	14:41	
Air Blank		0.000	14:41	
Control Test		0.078	14:42	
Air Blank		0.000	14:42	
Subject Sample	# 1	0.197	14:43	
Breath Volume		3.218	14:43	
Air Blank		0.000	14:43	
Air Blank		0.000	14:45	
Subject Sample	# 2	0.191	14:46	
Breath Volume		3.414	14:46	
Air Blank		0.000	14:46	
Control Test		0.077	14:47	
Air Blank			14:47	RFI Detect
Air Blank		0.000	14:47	

Data Download Date/Time: 01/31/2016 07:38

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ELLIS,ASHLEY
ARRESTING OFFICER AGENCY: Sunrise PD
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 01:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: FISHER

DATE/TIME OF TEST: 01/01/2016 02:03:53

Diagnostics Check		OK	01:57
Air Blank		0.000	01:58
Control Test		0.080	01:58
Air Blank		0.000	01:58
Subject Sample	# 1	0.185	01:59
Breath Volume		2.371	01:59
Air Blank		0.000	01:59
Air Blank		0.000	02:01
Subject Sample	# 2	0.186	02:02
Breath Volume		1.414	02:02
Air Blank		0.000	02:02
Control Test		0.080	02:03
Air Blank		0.000	02:03
Diagnostics Check		OK	02:03

Data Download Date/Time: 01/29/2016 18:45

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: MANNING,MARK
ARRESTING OFFICER AGENCY: Pembroke Pines PD
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:21

SUBJECT'S LAST NAME: CRANDALL

DATE/TIME OF TEST: 01/02/2016 03:50:27

Diagnostics Check		OK	03:43
Air Blank		0.000	03:43
Control Test		0.080	03:43
Air Blank		0.000	03:44
Subject Sample	# 1	0.041	03:45
Breath Volume		1.234	03:45
Air Blank		0.000	03:46
Air Blank		0.000	03:48
Subject Sample	# 2	0.042	03:48
Breath Volume		1.507	03:48
Air Blank		0.000	03:49
Control Test		0.080	03:49
Air Blank		0.000	03:50
Diagnostics Check		OK	03:50

Data Download Date/Time: 01/29/2016 18:45

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: CAROTTI,MOISES
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 01:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:16

SUBJECT'S LAST NAME: WILLIAMS JR

DATE/TIME OF TEST: 01/03/2016 02:43:26

Diagnostics Check		OK	02:36
Air Blank		0.000	02:36
Control Test		0.080	02:36
Air Blank		0.000	02:37
Subject Sample	# 1	0.117	02:38
Breath Volume		1.382	02:38
Air Blank		0.000	02:39
Air Blank		0.000	02:41
Subject Sample	# 2	0.122	02:41
Breath Volume		2.480	02:41
Air Blank		0.000	02:42
Control Test		0.079	02:42
Air Blank		0.000	02:43
Diagnostics Check		OK	02:43

Data Download Date/Time: 01/29/2016 18:45

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: CUSACK,JON
ARRESTING OFFICER AGENCY: Pembroke Pines PD
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 15:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:23

SUBJECT'S LAST NAME: SHAFEEK

DATE/TIME OF TEST: 01/03/2016 16:58:11

Diagnostics Check		OK	16:49
Air Blank		0.000	16:49
Control Test		0.080	16:50
Air Blank		0.000	16:50
Subject Sample	# 1	0.245	16:51
Breath Volume		1.648	16:51
Air Blank		0.000	16:51
Air Blank		0.000	16:53
Subject Sample	# 2	0.221	16:53
Breath Volume		1.660	16:53
Air Blank		0.000	16:54
Air Blank		0.000	16:56
Subject Sample	# 3	0.241	16:56
Breath Volume		1.566	16:56
Air Blank		0.000	16:57
Control Test		0.080	16:57
Air Blank		0.000	16:58
Diagnostics Check		OK	16:58

Data Download Date/Time: 01/29/2016 18:45

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: MERCK,CHRISTOPHER
ARRESTING OFFICER AGENCY: Lighthouse Point PD
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 14:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:16

SUBJECT'S LAST NAME: BURNS

DATE/TIME OF TEST: 01/03/2016 17:43:48

Diagnostics Check		OK	17:37
Air Blank		0.000	17:37
Control Test		0.080	17:38
Air Blank		0.000	17:38
Subject Sample	# 1	0.266	17:39
Breath Volume		1.683	17:39
Air Blank		0.000	17:39
Air Blank		0.000	17:41
Subject Sample	# 2	0.249	17:42
Breath Volume		2.156	17:42
Air Blank		0.000	17:42
Control Test		0.080	17:43
Air Blank		0.000	17:43
Diagnostics Check		OK	17:43

Data Download Date/Time: 01/29/2016 18:45

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: MANNING,MARK
ARRESTING OFFICER AGENCY: Pembroke Pines PD
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:47

SUBJECT'S LAST NAME: WATSON

DATE/TIME OF TEST: 01/04/2016 03:19:17

Diagnostics Check		OK	03:16	
Air Blank		0.000	03:16	
Control Test		0.080	03:17	
Air Blank		0.000	03:17	
Subject Sample	# 1		03:17	Subject Test Refused
Breath Volume		0.000	03:17	
Air Blank		0.000	03:18	
Control Test		0.080	03:18	
Air Blank		0.000	03:19	
Diagnostics Check		OK	03:19	

Data Download Date/Time: 01/29/2016 18:45

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: HARRIS,MILTON
ARRESTING OFFICER AGENCY: Pembroke Pines PD
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: LAPORTE

DATE/TIME OF TEST: 01/05/2016 00:47:23

Diagnostics Check		OK	00:41
Air Blank		0.000	00:41
Control Test		0.080	00:41
Air Blank		0.000	00:42
Subject Sample	# 1	0.241	00:42
Breath Volume		2.339	00:42
Air Blank		0.000	00:43
Air Blank		0.000	00:45
Subject Sample	# 2	0.235	00:45
Breath Volume		1.761	00:45
Air Blank		0.000	00:46
Control Test		0.080	00:46
Air Blank		0.000	00:47
Diagnostics Check		OK	00:47

Data Download Date/Time: 01/29/2016 18:45

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: BALUJA,JONATHAN
ARRESTING OFFICER AGENCY: Hollywood PD
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 03:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:22

SUBJECT'S LAST NAME: WATSON

DATE/TIME OF TEST: 01/08/2016 04:49:35

Diagnostics Check		OK	04:42
Air Blank		0.000	04:42
Control Test		0.079	04:43
Air Blank		0.000	04:43
Subject Sample	# 1	0.192	04:44
Breath Volume		1.367	04:44
Air Blank		0.000	04:45
Air Blank		0.000	04:47
Subject Sample	# 2	0.189	04:48
Breath Volume		1.523	04:48
Air Blank		0.000	04:48
Control Test		0.080	04:48
Air Blank		0.000	04:49
Diagnostics Check		OK	04:49

Data Download Date/Time: 01/29/2016 18:45

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARNANDAN,INDROWTIE,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SERPHOS,STEVEN
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 04:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:40

SUBJECT'S LAST NAME: GALLEGOS

DATE/TIME OF TEST: 01/09/2016 05:09:05

Diagnostics Check		OK	05:00
Air Blank		0.000	05:00
Control Test		0.081	05:01
Air Blank		0.000	05:01
Subject Sample	# 1	0.109	05:03
Breath Volume		1.421	05:03
Air Blank		0.000	05:04
Air Blank		0.000	05:06
Subject Sample	# 2	0.105	05:07
Breath Volume		1.378	05:07
Air Blank		0.000	05:08
Control Test		0.079	05:08
Air Blank		0.000	05:09
Diagnostics Check		OK	05:09

Data Download Date/Time: 01/29/2016 18:45

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUN,TONY,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: DANG,CHRISTOPHER
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 14:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:59

SUBJECT'S LAST NAME: MELVIN

DATE/TIME OF TEST: 01/09/2016 16:33:22

Diagnostics Check		OK	16:27
Air Blank		0.000	16:27
Control Test		0.081	16:28
Air Blank		0.000	16:28
Subject Sample	# 1	0.233	16:28
Breath Volume		1.324	16:28
Air Blank		0.000	16:29
Air Blank		0.000	16:31
Subject Sample	# 2	0.240	16:31
Breath Volume		2.449	16:31
Air Blank		0.000	16:32
Control Test		0.080	16:32
Air Blank		0.000	16:33
Diagnostics Check		OK	16:33

Data Download Date/Time: 01/29/2016 18:45

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINA,MARMILLY,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: IDARRAGA,CARLOS
ARRESTING OFFICER AGENCY: Hollywood PD
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 22:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: PAIGE

DATE/TIME OF TEST: 01/09/2016 23:20:49

Diagnostics Check		OK	23:11
Air Blank		0.000	23:11
Control Test		0.080	23:11
Air Blank		0.000	23:12
Subject Sample	# 1	0.239	23:12
Breath Volume		2.042	23:12
Air Blank		0.000	23:13
Air Blank		0.000	23:15
Subject Sample	# 2	0.214	23:16
Breath Volume		1.585	23:16
Air Blank		0.000	23:16
Air Blank		0.000	23:18
Subject Sample	# 3	0.219	23:19
Breath Volume		1.144	23:19
Air Blank		0.000	23:19
Control Test		0.079	23:20
Air Blank		0.000	23:20
Diagnostics Check		OK	23:20

Data Download Date/Time: 01/29/2016 18:45

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARNANDAN,INDROWTIE,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: HARRIS,MILTON
ARRESTING OFFICER AGENCY: Pembroke Pines PD
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 04:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:15

SUBJECT'S LAST NAME: OSORIO VALDIVIESO

DATE/TIME OF TEST: 01/14/2016 05:44:21

Diagnostics Check		OK	05:35
Air Blank		0.000	05:36
Control Test		0.079	05:36
Air Blank		0.000	05:37
Subject Sample	# 1	0.115	05:38
Breath Volume		1.234	05:38
Air Blank		0.000	05:39
Air Blank		0.000	05:40
Subject Sample	# 2	0.117	05:42
Breath Volume		1.667	05:42
Air Blank		0.000	05:43
Control Test		0.080	05:43
Air Blank		0.000	05:44
Diagnostics Check		OK	05:44

Data Download Date/Time: 01/29/2016 18:45

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: WASHINGTON,YVETTE
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 19:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:45

SUBJECT'S LAST NAME: MCNULTY

DATE/TIME OF TEST: 01/14/2016 20:12:36

Diagnostics Check		OK	20:05
Air Blank		0.000	20:05
Control Test		0.079	20:05
Air Blank		0.000	20:06
Subject Sample	# 1	0.366	20:07
Breath Volume		4.054	20:07
Air Blank		0.000	20:08
Air Blank		0.000	20:10
Subject Sample	# 2	0.368	20:11
Breath Volume		2.511	20:11
Air Blank		0.000	20:11
Control Test		0.079	20:12
Air Blank		0.000	20:12
Diagnostics Check		OK	20:12

Data Download Date/Time: 01/29/2016 18:45

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SPEZIALE,ADAM
ARRESTING OFFICER AGENCY: Margate PD
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 01:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: SAINT-JEAN

DATE/TIME OF TEST: 01/15/2016 02:35:56

Diagnostics Check		OK	02:29
Air Blank		0.000	02:29
Control Test		0.079	02:29
Air Blank		0.000	02:30
Subject Sample	# 1	0.087	02:31
Breath Volume		2.953	02:31
Air Blank		0.000	02:31
Air Blank		0.000	02:33
Subject Sample	# 2	0.080	02:34
Breath Volume		1.160	02:34
Air Blank		0.000	02:35
Control Test		0.080	02:35
Air Blank		0.000	02:35
Diagnostics Check		OK	02:35

Data Download Date/Time: 01/29/2016 18:45

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: WASHINGTON,YVETTE
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 04:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:37

SUBJECT'S LAST NAME: SOMOZA GARCIA

DATE/TIME OF TEST: 01/15/2016 05:10:41

Diagnostics Check		OK	05:00
Air Blank		0.000	05:00
Control Test		0.080	05:01
Air Blank		0.000	05:01
Subject Sample	# 1	0.167	05:03
Breath Volume		1.101	05:03
Air Blank		0.000	05:03
Air Blank		0.000	05:05
Subject Sample	# 2	0.192	05:06
Breath Volume		2.027	05:06
Air Blank		0.000	05:06
Air Blank		0.000	05:08
Subject Sample	# 3	0.169	05:09
Breath Volume		1.121	05:09
Air Blank		0.000	05:09
Control Test		0.080	05:10
Air Blank		0.000	05:10
Diagnostics Check		OK	05:10

Data Download Date/Time: 01/29/2016 18:45

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: LUNA,KRISTINA
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 19:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:23

SUBJECT'S LAST NAME: STEFFENS

DATE/TIME OF TEST: 01/15/2016 20:51:34

Diagnostics Check		OK	20:44
Air Blank		0.000	20:44
Control Test		0.080	20:45
Air Blank		0.000	20:45
Subject Sample	# 1	0.225	20:46
Breath Volume		3.390	20:46
Air Blank		0.000	20:47
Air Blank		0.000	20:49
Subject Sample	# 2	0.230	20:50
Breath Volume		4.750	20:50
Air Blank		0.000	20:50
Control Test		0.079	20:50
Air Blank		0.000	20:51
Diagnostics Check		OK	20:51

Data Download Date/Time: 01/29/2016 18:45

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SERPHOS,STEVEN
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 03:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:32

SUBJECT'S LAST NAME: FERRERA

DATE/TIME OF TEST: 01/16/2016 04:01:04

Diagnostics Check		OK	03:54
Air Blank		0.000	03:55
Control Test		0.080	03:55
Air Blank		0.000	03:56
Subject Sample	# 1	0.096	03:56
Breath Volume		2.773	03:56
Air Blank		0.000	03:57
Air Blank		0.000	03:59
Subject Sample	# 2	0.096	03:59
Breath Volume		2.968	03:59
Air Blank		0.000	04:00
Control Test		0.080	04:00
Air Blank		0.000	04:00
Diagnostics Check		OK	04:01

Data Download Date/Time: 01/29/2016 18:45

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: HUNEKE,DONNARD
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 00:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:42

SUBJECT'S LAST NAME: HAYEK

DATE/TIME OF TEST: 01/17/2016 01:09:49

Diagnostics Check		OK	01:03
Air Blank		0.000	01:03
Control Test		0.080	01:04
Air Blank		0.000	01:04
Subject Sample	# 1	0.187	01:05
Breath Volume		2.726	01:05
Air Blank		0.000	01:05
Air Blank		0.000	01:07
Subject Sample	# 2	0.181	01:08
Breath Volume		3.660	01:08
Air Blank		0.000	01:08
Control Test		0.079	01:09
Air Blank		0.000	01:09
Diagnostics Check		OK	01:09

Data Download Date/Time: 01/29/2016 18:45

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINA,MARMILLY,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: VELEZ ALICEA,YESENIA
ARRESTING OFFICER AGENCY: Hallandale Beach PD
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 02:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: SINGH

DATE/TIME OF TEST: 01/18/2016 03:00:01

Diagnostics Check		OK	02:53
Air Blank		0.000	02:53
Control Test		0.080	02:54
Air Blank		0.000	02:54
Subject Sample	# 1	0.169	02:55
Breath Volume		3.457	02:55
Air Blank		0.000	02:55
Air Blank		0.000	02:57
Subject Sample	# 2	0.173	02:58
Breath Volume		3.750	02:58
Air Blank		0.000	02:59
Control Test		0.079	02:59
Air Blank		0.000	02:59
Diagnostics Check		OK	03:00

Data Download Date/Time: 01/29/2016 18:45

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: MCCRAY,JOHNNY
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 00:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:48

SUBJECT'S LAST NAME: WILLIAMSON

DATE/TIME OF TEST: 01/19/2016 01:26:01

Diagnostics Check		OK	01:14	
Air Blank		0.000	01:14	
Control Test		0.080	01:15	
Air Blank		0.000	01:15	
Subject Sample	# 1	0.169	01:18	Volume Not Met
Breath Volume		0.082	01:18	
Air Blank		0.000	01:19	
Air Blank		0.000	01:21	
Subject Sample	# 2	0.186	01:24	Volume Not Met
Breath Volume		0.035	01:24	
Air Blank		0.000	01:25	
Control Test		0.079	01:25	
Air Blank		0.000	01:25	
Diagnostics Check		OK	01:26	

Data Download Date/Time: 01/29/2016 18:45

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: MCCRAY,JOHNNY
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 00:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:48

SUBJECT'S LAST NAME: WILLIAMSON

DATE/TIME OF TEST: 01/19/2016 01:38:55

Diagnostics Check		OK	01:31
Air Blank		0.000	01:31
Control Test		0.080	01:32
Air Blank		0.000	01:32
Subject Sample	# 1	0.179	01:34
Breath Volume		2.203	01:34
Air Blank		0.000	01:35
Air Blank		0.000	01:36
Subject Sample	# 2	0.176	01:37
Breath Volume		1.628	01:37
Air Blank		0.000	01:38
Control Test		0.080	01:38
Air Blank		0.000	01:38
Diagnostics Check		OK	01:38

Data Download Date/Time: 01/29/2016 18:45

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARNANDAN,INDROWTIE,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ETIENNE,DJOURY
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 01:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: CARVAJAL

DATE/TIME OF TEST: 01/20/2016 02:27:19

Diagnostics Check		OK	02:20
Air Blank		0.000	02:20
Control Test		0.080	02:21
Air Blank		0.000	02:21
Subject Sample	# 1	0.127	02:22
Breath Volume		2.042	02:22
Air Blank		0.000	02:23
Air Blank		0.000	02:25
Subject Sample	# 2	0.134	02:25
Breath Volume		2.226	02:25
Air Blank		0.000	02:26
Control Test		0.080	02:26
Air Blank		0.000	02:27
Diagnostics Check		OK	02:27

Data Download Date/Time: 01/29/2016 18:45

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SERPHOS,STEVEN
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 18:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:10

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 01/20/2016 19:42:07

Diagnostics Check		OK	19:35
Air Blank		0.000	19:35
Control Test		0.080	19:36
Air Blank		0.000	19:36
Subject Sample	# 1	0.155	19:37
Breath Volume		2.941	19:37
Air Blank		0.000	19:38
Air Blank		0.000	19:40
Subject Sample	# 2	0.172	19:40
Breath Volume		2.972	19:40
Air Blank		0.000	19:41
Control Test		0.080	19:41
Air Blank		0.000	19:42
Diagnostics Check		OK	19:42

Data Download Date/Time: 01/29/2016 18:45

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARNANDAN,INDROWTIE,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ROBERTS,CURTIS
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 04:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:43

SUBJECT'S LAST NAME: MELGAREJO

DATE/TIME OF TEST: 01/22/2016 05:09:47

Diagnostics Check		OK	05:03
Air Blank		0.000	05:04
Control Test		0.080	05:04
Air Blank		0.000	05:05
Subject Sample	# 1	0.000	05:05
Breath Volume		2.664	05:05
Air Blank		0.000	05:05
Air Blank		0.000	05:08
Subject Sample	# 2	0.000	05:08
Breath Volume		2.875	05:08
Air Blank		0.000	05:08
Control Test		0.080	05:09
Air Blank		0.000	05:09
Diagnostics Check		OK	05:09

Data Download Date/Time: 01/29/2016 18:45

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUN,TONY,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: WASHINGTON,YVETTE
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 01:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:44

SUBJECT'S LAST NAME: AVILA VELAZCO

DATE/TIME OF TEST: 01/23/2016 02:11:30

Diagnostics Check		OK	02:04
Air Blank		0.000	02:04
Control Test		0.080	02:05
Air Blank		0.000	02:05
Subject Sample	# 1	0.186	02:06
Breath Volume		1.527	02:06
Air Blank		0.000	02:07
Air Blank		0.000	02:09
Subject Sample	# 2	0.186	02:10
Breath Volume		1.488	02:10
Air Blank		0.000	02:10
Control Test		0.079	02:10
Air Blank		0.000	02:11
Diagnostics Check		OK	02:11

Data Download Date/Time: 01/29/2016 18:45

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: MUNOZ,BROWEL
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 07:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:18

SUBJECT'S LAST NAME: APONTE

DATE/TIME OF TEST: 01/23/2016 08:45:19

Diagnostics Check		OK	08:38
Air Blank		0.000	08:39
Control Test		0.081	08:39
Air Blank		0.000	08:40
Subject Sample	# 1	0.107	08:40
Breath Volume		1.742	08:40
Air Blank		0.000	08:41
Air Blank		0.000	08:43
Subject Sample	# 2	0.109	08:43
Breath Volume		2.972	08:43
Air Blank		0.000	08:44
Control Test		0.080	08:44
Air Blank		0.000	08:45
Diagnostics Check		OK	08:45

Data Download Date/Time: 01/29/2016 18:45

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARNANDAN,INDROWTIE,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: LAFLEUR,GREGORY
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 05:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:15

SUBJECT'S LAST NAME: DESAUGUSTE

DATE/TIME OF TEST: 01/27/2016 06:42:17

Diagnostics Check		OK	06:35
Air Blank		0.000	06:36
Control Test		0.079	06:36
Air Blank		0.000	06:36
Subject Sample	# 1	0.156	06:37
Breath Volume		1.828	06:37
Air Blank		0.000	06:38
Air Blank		0.000	06:40
Subject Sample	# 2	0.152	06:40
Breath Volume		1.300	06:40
Air Blank		0.000	06:41
Control Test		0.080	06:41
Air Blank		0.000	06:42
Diagnostics Check		OK	06:42

Data Download Date/Time: 01/29/2016 18:45

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DICIENZO,FRANCA,R
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: COFFIN,BRIAN
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 10:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:55

SUBJECT'S LAST NAME: NASH

DATE/TIME OF TEST: 01/27/2016 12:25:24

Diagnostics Check		OK	12:19
Air Blank		0.000	12:19
Control Test		0.080	12:19
Air Blank		0.000	12:20
Subject Sample	# 1	0.189	12:21
Breath Volume		1.835	12:21
Air Blank		0.000	12:21
Air Blank		0.000	12:23
Subject Sample	# 2	0.183	12:23
Breath Volume		2.378	12:23
Air Blank		0.000	12:24
Control Test		0.079	12:24
Air Blank		0.000	12:25
Diagnostics Check		OK	12:25

Data Download Date/Time: 01/29/2016 18:45

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SERPHOS,STEVEN
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 01:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: NEWMAN

DATE/TIME OF TEST: 01/28/2016 02:25:11

Diagnostics Check		OK	02:18
Air Blank		0.000	02:19
Control Test		0.080	02:19
Air Blank		0.000	02:20
Subject Sample	# 1	0.023	02:20
Breath Volume		3.570	02:20
Air Blank		0.000	02:21
Air Blank		0.000	02:23
Subject Sample	# 2	0.022	02:23
Breath Volume		3.367	02:23
Air Blank		0.000	02:24
Control Test		0.080	02:24
Air Blank		0.000	02:25
Diagnostics Check		OK	02:25

Data Download Date/Time: 01/29/2016 18:45

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ETIENNE,DJOURY
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 03:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:03

SUBJECT'S LAST NAME: GUZMAN

DATE/TIME OF TEST: 01/28/2016 04:30:41

Diagnostics Check		OK	04:23
Air Blank		0.000	04:23
Control Test		0.080	04:24
Air Blank		0.000	04:24
Subject Sample	# 1	0.119	04:26
Breath Volume		1.265	04:26
Air Blank		0.000	04:26
Air Blank		0.000	04:28
Subject Sample	# 2	0.111	04:29
Breath Volume		1.156	04:29
Air Blank		0.000	04:29
Control Test		0.079	04:30
Air Blank		0.000	04:30
Diagnostics Check		OK	04:30

Data Download Date/Time: 01/29/2016 18:45

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: STRZALKOWSKI,SEAN
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 22:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: ARGUETA

DATE/TIME OF TEST: 01/28/2016 23:07:19

Diagnostics Check		OK	23:00
Air Blank		0.000	23:00
Control Test		0.080	23:01
Air Blank		0.000	23:01
Subject Sample	# 1	0.211	23:02
Breath Volume		2.628	23:02
Air Blank		0.000	23:03
Air Blank		0.000	23:05
Subject Sample	# 2	0.203	23:05
Breath Volume		1.945	23:05
Air Blank		0.000	23:06
Control Test		0.080	23:06
Air Blank		0.000	23:07
Diagnostics Check		OK	23:07

Data Download Date/Time: 01/29/2016 18:45

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SAPP,JOSHUA
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 01:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:37

SUBJECT'S LAST NAME: METCALF

DATE/TIME OF TEST: 01/29/2016 02:06:42

Diagnostics Check		OK	02:00
Air Blank		0.000	02:01
Control Test		0.081	02:01
Air Blank		0.000	02:01
Subject Sample	# 1	0.218	02:02
Breath Volume		2.144	02:02
Air Blank		0.000	02:02
Air Blank		0.000	02:04
Subject Sample	# 2	0.215	02:05
Breath Volume		2.132	02:05
Air Blank		0.000	02:05
Control Test		0.080	02:06
Air Blank		0.000	02:06
Diagnostics Check		OK	02:06

Data Download Date/Time: 01/29/2016 18:45

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: HUNEKE,DONNARD
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 03:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:15

SUBJECT'S LAST NAME: TORPY

DATE/TIME OF TEST: 01/29/2016 04:42:02

Diagnostics Check		OK	04:35
Air Blank		0.000	04:35
Control Test		0.079	04:35
Air Blank		0.000	04:36
Subject Sample	# 1	0.182	04:37
Breath Volume		2.136	04:37
Air Blank		0.000	04:38
Air Blank		0.000	04:40
Subject Sample	# 2	0.176	04:40
Breath Volume		2.312	04:40
Air Blank		0.000	04:41
Control Test		0.079	04:41
Air Blank		0.000	04:41
Diagnostics Check		OK	04:42

Data Download Date/Time: 01/29/2016 18:45

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRINKMAN,KELLY,N
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: BRINKMAN,KELLY
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:53

SUBJECT'S LAST NAME: MONTERROSA ALFARO

DATE/TIME OF TEST: 01/01/2016 22:20:14

Diagnostics Check		OK	22:13
Air Blank		0.000	22:13
Control Test		0.080	22:14
Air Blank		0.000	22:14
Subject Sample	# 1	0.113	22:15
Breath Volume		1.214	22:15
Air Blank		0.000	22:16
Air Blank		0.000	22:18
Subject Sample	# 2	0.114	22:18
Breath Volume		2.019	22:18
Air Blank		0.000	22:19
Control Test		0.080	22:19
Air Blank		0.000	22:20
Diagnostics Check		OK	22:20

Data Download Date/Time: 01/12/2016 11:47

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ENGLAND,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: ENGLAND,CHRISTOPHER
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:53

SUBJECT'S LAST NAME: STAFFORD

DATE/TIME OF TEST: 01/01/2016 23:23:51

Diagnostics Check		OK	23:14	
Air Blank		0.000	23:14	
Control Test		0.079	23:14	
Air Blank		0.000	23:15	
Subject Sample	# 1		23:18	No Sample Provided
Breath Volume		0.000	23:18	
Air Blank		0.000	23:19	
Air Blank		0.000	23:21	
Subject Sample	# 2		23:22	Subject Test Refused
Breath Volume		0.000	23:22	
Air Blank		0.000	23:22	
Control Test		0.080	23:23	
Air Blank		0.000	23:23	
Diagnostics Check		OK	23:23	

Data Download Date/Time: 01/12/2016 11:47

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRINKMAN,KELLY,N
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: BRINKMAN,KELLY
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: GOTTA

DATE/TIME OF TEST: 01/03/2016 01:02:05

Diagnostics Check		OK	00:55
Air Blank		0.000	00:55
Control Test		0.079	00:56
Air Blank		0.000	00:56
Subject Sample	# 1	0.095	00:57
Breath Volume		1.164	00:57
Air Blank		0.000	00:57
Air Blank		0.000	00:59
Subject Sample	# 2	0.104	01:00
Breath Volume		1.675	01:00
Air Blank		0.000	01:01
Control Test		0.080	01:01
Air Blank		0.000	01:01
Diagnostics Check		OK	01:02

Data Download Date/Time: 01/12/2016 11:47

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICHMOND,MATTHEW,D
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: RICHMOND,MATTHEW
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 16:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:10

SUBJECT'S LAST NAME: LARTZ 3

DATE/TIME OF TEST: 01/04/2016 17:39:51

Diagnostics Check		OK	17:32
Air Blank		0.000	17:33
Control Test		0.080	17:33
Air Blank		0.000	17:33
Subject Sample	# 1	0.165	17:35
Breath Volume		3.585	17:35
Air Blank		0.000	17:35
Air Blank		0.000	17:37
Subject Sample	# 2	0.160	17:38
Breath Volume		3.410	17:38
Air Blank		0.000	17:38
Control Test		0.080	17:39
Air Blank		0.000	17:39
Diagnostics Check		OK	17:39

Data Download Date/Time: 01/12/2016 11:47

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ENGLAND,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: ENGLAND,CHRISTOPHER
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:24

SUBJECT'S LAST NAME: TOLLIVER

DATE/TIME OF TEST: 01/08/2016 05:01:04

Diagnostics Check		OK	04:44	
Air Blank		0.000	04:45	
Control Test		0.079	04:45	
Air Blank		0.000	04:45	
Subject Sample	# 1	0.073	04:49	Volume Not Met
Breath Volume		0.097	04:49	
Air Blank		0.000	04:49	
Air Blank		0.000	04:51	
Subject Sample	# 2	0.075	04:54	
Breath Volume		2.316	04:54	
Air Blank		0.000	04:54	
Air Blank		0.000	04:56	
Subject Sample	# 3		04:59	No Sample Provided
Breath Volume		0.000	04:59	
Air Blank		0.000	05:00	
Control Test		0.077	05:00	
Air Blank		0.000	05:00	
Diagnostics Check		OK	05:01	

Data Download Date/Time: 01/12/2016 11:47

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LODER,D,J
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: WHITE,C
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: STEPHENS

DATE/TIME OF TEST: 01/10/2016 01:50:05

Diagnostics Check		OK	01:42
Air Blank		0.000	01:43
Control Test		0.080	01:43
Air Blank		0.000	01:44
Subject Sample	# 1	0.243	01:45
Breath Volume		1.523	01:45
Air Blank		0.000	01:46
Air Blank		0.000	01:47
Subject Sample	# 2	0.242	01:48
Breath Volume		1.621	01:48
Air Blank		0.000	01:49
Control Test		0.078	01:49
Air Blank		0.000	01:50
Diagnostics Check		OK	01:50

Data Download Date/Time: 01/12/2016 11:47

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ENGLAND,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: ENGLAND,CHRISTOPHER
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: GALARZA-PALMA

DATE/TIME OF TEST: 01/10/2016 03:03:48

Diagnostics Check		OK	02:57	
Air Blank		0.000	02:57	
Control Test		0.078	02:57	
Air Blank		0.000	02:58	
Subject Sample	# 1	0.106	02:58	
Breath Volume		1.253	02:58	
Air Blank		0.000	02:59	
Air Blank		0.000	03:01	
Subject Sample	# 2	0.111	03:02	
Breath Volume		3.035	03:02	
Air Blank			03:03	Purge Fail
Air Blank		0.000	03:03	

Data Download Date/Time: 01/12/2016 11:47

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ENGLAND,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: ENGLAND,CHRISTOPHER
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: GALARZA-PALMA

DATE/TIME OF TEST: 01/10/2016 03:07:37

Diagnostics Check	OK	03:06	
Air Blank	0.000	03:06	
Control Test	0.069	03:07	Control Outside Tolerance
Air Blank	0.000	03:07	

Data Download Date/Time: 01/12/2016 11:47

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ENGLAND,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: ENGLAND,CHRISTOPHER
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 12/11/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: GALARZA-PALMA

DATE/TIME OF TEST: 01/10/2016 03:14:16

Diagnostics Check	OK	03:12	
Air Blank	0.000	03:13	
Control Test	0.072	03:13	Control Outside Tolerance
Air Blank	0.000	03:14	

Data Download Date/Time: 01/12/2016 11:47

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON,ADRIAN,P
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: WASHINGTON,ADRIAN
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 01/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 17:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:34

SUBJECT'S LAST NAME: MOORE

DATE/TIME OF TEST: 01/12/2016 18:01:51

Diagnostics Check		OK	17:54
Air Blank		0.000	17:55
Control Test		0.079	17:55
Air Blank		0.000	17:56
Subject Sample	# 1	0.119	17:57
Breath Volume		2.804	17:57
Air Blank		0.000	17:58
Air Blank		0.000	17:59
Subject Sample	# 2	0.117	18:00
Breath Volume		3.019	18:00
Air Blank		0.000	18:00
Control Test		0.078	18:01
Air Blank		0.000	18:01
Diagnostics Check		OK	18:01

Data Download Date/Time: 02/04/2016 11:47

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WATTS,ELIZABETH,P
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: BALMER,PAUL
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 01/12/2016

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:22

SUBJECT'S LAST NAME: NOXON

DATE/TIME OF TEST: 01/14/2016 21:57:45

Diagnostics Check		OK	21:45	
Air Blank		0.000	21:46	
Control Test		0.080	21:46	
Air Blank		0.000	21:47	
Subject Sample	# 1		21:50	No Sample Provided
Breath Volume		0.000	21:50	
Air Blank		0.000	21:51	
Air Blank		0.000	21:52	
Subject Sample	# 2	0.173	21:55	Volume Not Met
Breath Volume		0.746	21:55	
Air Blank		0.000	21:56	
Control Test		0.075	21:57	
Air Blank		0.000	21:57	
Diagnostics Check		OK	21:57	

Data Download Date/Time: 02/04/2016 11:47

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NUGENT,DAVID,A
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: NUGENT,DAVID
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 01/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: WHITESELL

DATE/TIME OF TEST: 01/17/2016 02:07:32

Diagnostics Check		OK	02:01
Air Blank		0.000	02:01
Control Test		0.079	02:01
Air Blank		0.000	02:02
Subject Sample	# 1	0.197	02:02
Breath Volume		2.746	02:02
Air Blank		0.000	02:03
Air Blank		0.000	02:05
Subject Sample	# 2	0.182	02:05
Breath Volume		1.820	02:05
Air Blank		0.000	02:06
Control Test		0.077	02:06
Air Blank		0.000	02:07
Diagnostics Check		OK	02:07

Data Download Date/Time: 02/04/2016 11:47

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRINKMAN,KELLY,
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: BRINKMAN,KELLY
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 01/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:29

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 01/22/2016 03:57:06

Diagnostics Check		OK	03:50
Air Blank		0.000	03:50
Control Test		0.079	03:51
Air Blank		0.000	03:51
Subject Sample	# 1	0.168	03:52
Breath Volume		3.531	03:52
Air Blank		0.000	03:53
Air Blank		0.000	03:54
Subject Sample	# 2	0.161	03:55
Breath Volume		3.167	03:55
Air Blank		0.000	03:56
Control Test		0.080	03:56
Air Blank		0.000	03:57
Diagnostics Check		OK	03:57

Data Download Date/Time: 02/04/2016 11:47

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ENGLAND,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: ENGLAND,CHRISTOPHER
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 01/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: MILLAN

DATE/TIME OF TEST: 01/29/2016 01:32:46

Diagnostics Check		OK	01:24
Air Blank		0.000	01:25
Control Test		0.080	01:25
Air Blank		0.000	01:26
Subject Sample	# 1	0.161	01:28
Breath Volume		1.847	01:28
Air Blank		0.000	01:29
Air Blank		0.000	01:30
Subject Sample	# 2	0.166	01:31
Breath Volume		2.246	01:31
Air Blank		0.000	01:31
Control Test		0.081	01:32
Air Blank		0.000	01:32
Diagnostics Check		OK	01:32

Data Download Date/Time: 02/04/2016 11:47

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ENGLAND,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: ENGLAND,CHRISTOPHER
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 01/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:23

SUBJECT'S LAST NAME: PITTMAN

DATE/TIME OF TEST: 01/29/2016 03:54:31

Diagnostics Check		OK	03:44	
Air Blank		0.000	03:45	
Control Test		0.079	03:45	
Air Blank		0.000	03:45	
Subject Sample	# 1		03:49	No Sample Provided
Breath Volume		0.000	03:49	
Air Blank		0.000	03:49	
Air Blank		0.000	03:51	
Subject Sample	# 2		03:53	Subject Test Refused
Breath Volume		0.000	03:53	
Air Blank		0.000	03:53	
Control Test		0.078	03:53	
Air Blank		0.000	03:54	
Diagnostics Check		OK	03:54	

Data Download Date/Time: 02/04/2016 11:47

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001071

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALTSTATT,MATTHEW,B
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: SAUNDERS,J
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 12/21/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28814080A3
DRY GAS EXPIRY DATE: 11/05/2016
VIOLATION CODE: DUI
ARREST TIME: 01:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:04

SUBJECT'S LAST NAME: GARZA

DATE/TIME OF TEST: 01/01/2016 02:34:57

Diagnostics Check		OK	02:28
Air Blank		0.000	02:28
Control Test		0.082	02:28
Air Blank		0.000	02:29
Subject Sample	# 1	0.140	02:29
Breath Volume		3.375	02:29
Air Blank		0.000	02:30
Air Blank		0.000	02:32
Subject Sample	# 2	0.133	02:33
Breath Volume		3.386	02:33
Air Blank		0.000	02:34
Control Test		0.080	02:34
Air Blank		0.000	02:34
Diagnostics Check		OK	02:34

Data Download Date/Time: 01/21/2016 20:35

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001071

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,DEPUTY,J
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: CRUZ,OFFICER
ARRESTING OFFICER AGENCY: Orange Park PD
LAST AGENCY INSPECTION: 12/21/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28814080A3
DRY GAS EXPIRY DATE: 11/05/2016
VIOLATION CODE: DUI
ARREST TIME: 00:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: DUPLESSY

DATE/TIME OF TEST: 01/08/2016 04:30:29

Diagnostics Check		OK	04:24
Air Blank		0.000	04:24
Control Test		0.081	04:24
Air Blank		0.000	04:25
Subject Sample	# 1	0.000	04:25
Breath Volume		1.527	04:25
Air Blank		0.000	04:26
Air Blank		0.000	04:28
Subject Sample	# 2	0.000	04:29
Breath Volume		2.832	04:29
Air Blank		0.000	04:29
Control Test		0.081	04:29
Air Blank		0.000	04:30
Diagnostics Check		OK	04:30

Data Download Date/Time: 01/21/2016 20:35

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001071

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALTSTATT,MATTHEW,B
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: LOVETT,T
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 12/21/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28814080A3
DRY GAS EXPIRY DATE: 11/05/2016
VIOLATION CODE: DUI
ARREST TIME: 03:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: HOWARD

DATE/TIME OF TEST: 01/14/2016 03:58:56

Diagnostics Check		OK	03:47	
Air Blank		0.000	03:47	
Control Test		0.081	03:48	
Air Blank		0.000	03:48	
Subject Sample	# 1	0.162	03:51	Volume Not Met
Breath Volume		0.000	03:51	
Air Blank		0.000	03:52	
Air Blank		0.000	03:54	
Subject Sample	# 2	0.207	03:57	Volume Not Met
Breath Volume		0.210	03:57	
Air Blank		0.000	03:58	
Control Test		0.081	03:58	
Air Blank		0.000	03:58	
Diagnostics Check		OK	03:58	

Data Download Date/Time: 01/21/2016 20:35

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001071

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEBLANC,C,A
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: PRITCHARD,T
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 12/21/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28814080A3
DRY GAS EXPIRY DATE: 11/05/2016
VIOLATION CODE: DUI
ARREST TIME: 16:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:23

SUBJECT'S LAST NAME: BESSENT

DATE/TIME OF TEST: 01/15/2016 17:50:25

Diagnostics Check		OK	17:44
Air Blank		0.000	17:44
Control Test		0.081	17:44
Air Blank		0.000	17:45
Subject Sample	# 1	0.000	17:45
Breath Volume		1.160	17:45
Air Blank		0.000	17:46
Air Blank		0.000	17:48
Subject Sample	# 2	0.000	17:49
Breath Volume		2.097	17:49
Air Blank		0.000	17:49
Control Test		0.081	17:49
Air Blank		0.000	17:50
Diagnostics Check		OK	17:50

Data Download Date/Time: 01/21/2016 20:35

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001071

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DURRETTE,T,K
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: MASON,M
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 12/21/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28814080A3
DRY GAS EXPIRY DATE: 11/05/2016
VIOLATION CODE: DUI
ARREST TIME: 11:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:10

SUBJECT'S LAST NAME: MCDANIEL

DATE/TIME OF TEST: 01/18/2016 12:38:00

Diagnostics Check		OK	12:31
Air Blank		0.000	12:31
Control Test		0.083	12:32
Air Blank		0.000	12:32
Subject Sample	# 1	0.000	12:33
Breath Volume		2.820	12:33
Air Blank		0.000	12:33
Air Blank		0.000	12:35
Subject Sample	# 2	0.000	12:36
Breath Volume		1.750	12:36
Air Blank		0.000	12:37
Control Test		0.082	12:37
Air Blank		0.000	12:37
Diagnostics Check		OK	12:38

Data Download Date/Time: 01/21/2016 20:35

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001071

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,DEPUTY,J
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: LLEWELLYN,DEPUTY
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 01/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28814080A3
DRY GAS EXPIRY DATE: 11/05/2016
VIOLATION CODE: DUI
ARREST TIME: 20:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: NOLAN

DATE/TIME OF TEST: 01/21/2016 21:21:52

Diagnostics Check		OK	21:15
Air Blank		0.000	21:16
Control Test		0.081	21:16
Air Blank		0.000	21:16
Subject Sample	# 1	0.026	21:17
Breath Volume		3.164	21:17
Air Blank		0.000	21:17
Air Blank		0.000	21:19
Subject Sample	# 2	0.024	21:20
Breath Volume		2.140	21:20
Air Blank		0.000	21:20
Control Test		0.081	21:21
Air Blank		0.000	21:21
Diagnostics Check		OK	21:21

Data Download Date/Time: 02/21/2016 18:11

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001071

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZIER,TAMMY,L
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: VOLLER,DAVID
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 01/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28814080A3
DRY GAS EXPIRY DATE: 11/05/2016
VIOLATION CODE: DUI
ARREST TIME: 01:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:36

SUBJECT'S LAST NAME: HAYDU

DATE/TIME OF TEST: 01/30/2016 02:11:01

Diagnostics Check		OK	02:02	
Air Blank		0.000	02:03	
Control Test		0.081	02:03	
Air Blank		0.000	02:03	
Subject Sample	# 1		02:07	No Sample Provided
Breath Volume		0.000	02:07	
Air Blank		0.000	02:07	
Air Blank		0.000	02:09	
Subject Sample	# 2		02:09	Subject Test Refused
Breath Volume		0.000	02:09	
Air Blank		0.000	02:10	
Control Test		0.081	02:10	
Air Blank		0.000	02:10	
Diagnostics Check		OK	02:11	

Data Download Date/Time: 02/21/2016 18:11

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001071

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZIER,TAMMY,L
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: LINEBERRY,SCOTT
ARRESTING OFFICER AGENCY: Orange Park PD
LAST AGENCY INSPECTION: 01/21/2016

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 28814080A3
DRY GAS EXPIRY DATE: 11/05/2016
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:51

SUBJECT'S LAST NAME: CURTISS

DATE/TIME OF TEST: 01/30/2016 04:25:04

Diagnostics Check		OK	04:18
Air Blank		0.000	04:18
Control Test		0.082	04:19
Air Blank		0.000	04:19
Subject Sample	# 1	0.119	04:20
Breath Volume		2.109	04:20
Air Blank		0.000	04:21
Air Blank		0.000	04:23
Subject Sample	# 2	0.116	04:23
Breath Volume		1.867	04:23
Air Blank		0.000	04:24
Control Test		0.081	04:24
Air Blank		0.000	04:24
Diagnostics Check		OK	04:25

Data Download Date/Time: 02/21/2016 18:11

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001071

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAUNDERS,DEPUTY,C
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: PRITCHARD,DEPUTY
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 01/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28814080A3
DRY GAS EXPIRY DATE: 11/05/2016
VIOLATION CODE: DUI
ARREST TIME: 03:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:20

SUBJECT'S LAST NAME: COTTOM

DATE/TIME OF TEST: 01/31/2016 07:54:13

Diagnostics Check		OK	07:47
Air Blank		0.000	07:48
Control Test		0.082	07:48
Air Blank		0.000	07:49
Subject Sample	# 1	0.069	07:49
Breath Volume		1.496	07:49
Air Blank		0.000	07:50
Air Blank		0.000	07:52
Subject Sample	# 2	0.063	07:52
Breath Volume		1.132	07:52
Air Blank		0.000	07:53
Control Test		0.082	07:53
Air Blank		0.000	07:54
Diagnostics Check		OK	07:54

Data Download Date/Time: 02/21/2016 18:11

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001071

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PATRONE,RICHARD,L
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: BRIGGS,G
ARRESTING OFFICER AGENCY: Orange Park PD
LAST AGENCY INSPECTION: 01/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28814080A3
DRY GAS EXPIRY DATE: 11/05/2016
VIOLATION CODE: DUI
ARREST TIME: 22:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: CERCY

DATE/TIME OF TEST: 01/31/2016 23:14:57

Diagnostics Check		OK	23:06	
Air Blank		0.000	23:06	
Control Test		0.081	23:07	
Air Blank		0.000	23:07	
Subject Sample	# 1		23:10	No Sample Provided
Breath Volume		0.000	23:10	
Air Blank		0.000	23:11	
Air Blank		0.000	23:13	
Subject Sample	# 2		23:13	Subject Test Refused
Breath Volume		0.000	23:13	
Air Blank		0.000	23:14	
Control Test		0.081	23:14	
Air Blank		0.000	23:14	
Diagnostics Check		OK	23:14	

Data Download Date/Time: 02/21/2016 18:11

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001075

INSTRUMENT REGISTERED TO: FHP TROOP E
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SERRANO,MARILYN,
BREATH TEST OPERATOR AGENCY: FHP Troop K
ARRESTING OFFICER: SERRANO,MARILYN
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 12/03/2015

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1824329
DRY GAS EXPIRY DATE: 03/09/2018
VIOLATION CODE: DUI
ARREST TIME: 10:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:46

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 01/21/2016 00:13:21

Diagnostics Check		OK	00:10	
Air Blank		0.000	00:10	
Control Test		0.080	00:11	
Air Blank		0.000	00:11	
Subject Sample	# 1		00:11	Subject Test Refused
Breath Volume		0.000	00:11	
Air Blank		0.000	00:12	
Control Test		0.080	00:12	
Air Blank		0.000	00:13	
Diagnostics Check		OK	00:13	

Data Download Date/Time: 01/22/2016 09:34

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001075

INSTRUMENT REGISTERED TO: FHP TROOP E
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SERRANO,MARILYN,
BREATH TEST OPERATOR AGENCY: FHP Troop K
ARRESTING OFFICER: SERRANO,MARILYN
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 01/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1824329
DRY GAS EXPIRY DATE: 03/09/2018
VIOLATION CODE: DUI
ARREST TIME: 11:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:26

SUBJECT'S LAST NAME: BERGESIO

DATE/TIME OF TEST: 01/27/2016 01:07:55

Diagnostics Check		OK	00:55	
Air Blank		0.000	00:56	
Control Test		0.080	00:56	
Air Blank		0.000	00:57	
Subject Sample	# 1	0.154	01:00	Volume Not Met
Breath Volume		0.050	01:00	
Air Blank		0.000	01:00	
Air Blank		0.000	01:02	
Subject Sample	# 2	0.189	01:03	
Breath Volume		2.710	01:03	
Air Blank		0.000	01:03	
Air Blank		0.000	01:05	
Subject Sample	# 3	0.182	01:06	
Breath Volume		1.656	01:06	
Air Blank		0.000	01:06	
Control Test		0.079	01:07	
Air Blank		0.000	01:07	
Diagnostics Check		OK	01:07	

Data Download Date/Time: 02/27/2016 12:49

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001075

INSTRUMENT REGISTERED TO: FHP TROOP E
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADDISON,RAYMOND,E
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: CORTES,HECTOR
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 01/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1824329
DRY GAS EXPIRY DATE: 03/09/2018
VIOLATION CODE: DUI
ARREST TIME: 05:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:55

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 01/31/2016 06:28:33

Diagnostics Check		OK	06:19
Air Blank		0.000	06:20
Control Test		0.080	06:20
Air Blank		0.000	06:21
Subject Sample	# 1	0.153	06:23
Breath Volume		2.468	06:23
Air Blank		0.000	06:24
Air Blank		0.000	06:26
Subject Sample	# 2	0.155	06:27
Breath Volume		2.156	06:27
Air Blank		0.000	06:27
Control Test		0.079	06:27
Air Blank		0.000	06:28
Diagnostics Check		OK	06:28

Data Download Date/Time: 02/27/2016 12:49

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001075

INSTRUMENT REGISTERED TO: FHP TROOP E
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADDISON,RAYMOND,E
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: CORTES,HECTOR
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 01/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1824329
DRY GAS EXPIRY DATE: 03/09/2018
VIOLATION CODE: DUI
ARREST TIME: 05:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:27

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 01/31/2016 06:55:37

Diagnostics Check		OK	06:48
Air Blank		0.000	06:49
Control Test		0.080	06:49
Air Blank		0.000	06:50
Subject Sample	# 1	0.142	06:50
Breath Volume		2.015	06:50
Air Blank		0.000	06:51
Air Blank		0.000	06:53
Subject Sample	# 2	0.145	06:54
Breath Volume		2.371	06:54
Air Blank		0.000	06:54
Control Test		0.080	06:55
Air Blank		0.000	06:55
Diagnostics Check		OK	06:55

Data Download Date/Time: 02/27/2016 12:49

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001076

INSTRUMENT REGISTERED TO: ZEPHYRHILLS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARDNER,NATHAN,R
BREATH TEST OPERATOR AGENCY: Zephyrhills PD
ARRESTING OFFICER: UPCHURCH,MORGAN
ARRESTING OFFICER AGENCY: Zephyrhills PD
LAST AGENCY INSPECTION: 12/20/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07515080A4
DRY GAS EXPIRY DATE: 04/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:47

SUBJECT'S LAST NAME: CARTELAMI

DATE/TIME OF TEST: 01/08/2016 00:18:08

Diagnostics Check		OK	00:12
Air Blank		0.000	00:12
Control Test		0.077	00:12
Air Blank		0.000	00:13
Subject Sample	# 1	0.275	00:13
Breath Volume		1.816	00:13
Air Blank		0.000	00:14
Air Blank		0.000	00:16
Subject Sample	# 2	0.260	00:16
Breath Volume		1.242	00:16
Air Blank		0.000	00:17
Control Test		0.078	00:17
Air Blank		0.000	00:18
Diagnostics Check		OK	00:18

Data Download Date/Time: 01/28/2016 19:09

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001077

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POWELL,BRIAN,K
BREATH TEST OPERATOR AGENCY: Hardee County SO
ARRESTING OFFICER: CRUZ,IRVIN DAVID
ARRESTING OFFICER AGENCY: Hardee County SO
LAST AGENCY INSPECTION: 12/29/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1720515
DRY GAS EXPIRY DATE: 07/31/2017
VIOLATION CODE: DUI
ARREST TIME: 22:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:08

SUBJECT'S LAST NAME: VASQUEZ

DATE/TIME OF TEST: 01/16/2016 23:35:10

Diagnostics Check		OK	23:28
Air Blank		0.000	23:29
Control Test		0.078	23:29
Air Blank		0.000	23:29
Subject Sample	# 1	0.220	23:30
Breath Volume		2.542	23:30
Air Blank		0.000	23:31
Air Blank		0.000	23:33
Subject Sample	# 2	0.216	23:33
Breath Volume		2.457	23:33
Air Blank		0.000	23:34
Control Test		0.078	23:34
Air Blank		0.000	23:35
Diagnostics Check		OK	23:35

Data Download Date/Time: 01/26/2016 11:26

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001077

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONEY,MARY,L
BREATH TEST OPERATOR AGENCY: Hardee County SO
ARRESTING OFFICER: DAVIS,MIKE
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 12/29/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1720515
DRY GAS EXPIRY DATE: 07/31/2017
VIOLATION CODE: DUI
ARREST TIME: 21:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:48

SUBJECT'S LAST NAME: BECK

DATE/TIME OF TEST: 01/24/2016 23:19:43

Diagnostics Check		OK	23:08	
Air Blank		0.000	23:09	
Control Test		0.079	23:09	
Air Blank		0.000	23:10	
Subject Sample	# 1		23:13	No Sample Provided
Breath Volume		0.000	23:13	
Air Blank		0.000	23:13	
Air Blank		0.000	23:15	
Subject Sample	# 2		23:18	RFI Detect
Breath Volume		0.339	23:18	
Air Blank		0.000	23:18	
Control Test		0.079	23:19	
Air Blank		0.000	23:19	
Diagnostics Check		OK	23:19	

Data Download Date/Time: 01/26/2016 11:26

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001077

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONEY,MARY,L
BREATH TEST OPERATOR AGENCY: Hardee County SO
ARRESTING OFFICER: DAVIS,MIKE
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 12/29/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1720515
DRY GAS EXPIRY DATE: 07/31/2017
VIOLATION CODE: DUI
ARREST TIME: 21:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:48

SUBJECT'S LAST NAME: BECK

DATE/TIME OF TEST: 01/24/2016 23:38:35

Diagnostics Check		OK	23:23	
Air Blank		0.000	23:23	
Control Test		0.078	23:24	
Air Blank		0.000	23:24	
Subject Sample	# 1	0.133	23:25	
Breath Volume		1.523	23:25	
Air Blank		0.000	23:26	
Air Blank		0.000	23:28	
Subject Sample	# 2	0.117	23:31	Volume Not Met
Breath Volume		0.343	23:31	
Air Blank		0.000	23:31	
Air Blank		0.000	23:33	
Subject Sample	# 3	0.114	23:36	Volume Not Met
Breath Volume		0.148	23:36	
Air Blank		0.000	23:37	
Control Test		0.076	23:37	
Air Blank		0.000	23:38	
Diagnostics Check		OK	23:38	

Data Download Date/Time: 01/26/2016 11:26

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001077

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CONKLIN,ANTHONY,D
BREATH TEST OPERATOR AGENCY: Hardee County SO
ARRESTING OFFICER: ALEXANDER,DAVID
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 01/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1720515
DRY GAS EXPIRY DATE: 07/31/2017
VIOLATION CODE: DUI
ARREST TIME: 05:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:55

SUBJECT'S LAST NAME: PONCE

DATE/TIME OF TEST: 01/31/2016 06:21:59

Diagnostics Check		OK	06:15
Air Blank		0.000	06:16
Control Test		0.078	06:16
Air Blank		0.000	06:17
Subject Sample	# 1	0.074	06:17
Breath Volume		2.480	06:17
Air Blank		0.000	06:18
Air Blank		0.000	06:20
Subject Sample	# 2	0.075	06:20
Breath Volume		2.820	06:20
Air Blank		0.000	06:21
Control Test		0.078	06:21
Air Blank		0.000	06:21
Diagnostics Check		OK	06:21

Data Download Date/Time: 02/23/2016 12:41

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTER,MATTHEW,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ARKOVICH,ROBBIE
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/25/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 03:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: CANNON

DATE/TIME OF TEST: 01/01/2016 03:34:59

Diagnostics Check		OK	03:31	
Air Blank		0.000	03:31	
Control Test		0.079	03:32	
Air Blank		0.000	03:32	
Subject Sample	# 1		03:33	Subject Test Refused
Breath Volume		0.000	03:33	
Air Blank		0.000	03:34	
Control Test		0.078	03:34	
Air Blank		0.000	03:34	
Diagnostics Check		OK	03:34	

Data Download Date/Time: 01/30/2016 15:13

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NAGLE,TERRI,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ARKOVICH,ROBBIE
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/25/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 04:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:25

SUBJECT'S LAST NAME: LETZRING

DATE/TIME OF TEST: 01/01/2016 04:51:41

Diagnostics Check		OK	04:45
Air Blank		0.000	04:45
Control Test		0.078	04:46
Air Blank		0.000	04:46
Subject Sample	# 1	0.132	04:47
Breath Volume		2.769	04:47
Air Blank		0.000	04:47
Air Blank		0.000	04:49
Subject Sample	# 2	0.122	04:50
Breath Volume		2.398	04:50
Air Blank		0.000	04:50
Control Test		0.078	04:51
Air Blank		0.000	04:51
Diagnostics Check		OK	04:51

Data Download Date/Time: 01/30/2016 15:13

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTER,MATTHEW,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ARKOVICH,ROBBIE
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/25/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 01:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:12

SUBJECT'S LAST NAME: FERNANDEZ

DATE/TIME OF TEST: 01/07/2016 01:43:01

Diagnostics Check		OK	01:35
Air Blank		0.000	01:35
Control Test		0.079	01:35
Air Blank		0.000	01:36
Subject Sample	# 1	0.178	01:38
Breath Volume		1.765	01:38
Air Blank		0.000	01:39
Air Blank		0.000	01:40
Subject Sample	# 2	0.175	01:41
Breath Volume		1.332	01:41
Air Blank		0.000	01:42
Control Test		0.079	01:42
Air Blank		0.000	01:42
Diagnostics Check		OK	01:43

Data Download Date/Time: 01/30/2016 15:13

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARKOVICH,ROBBIE,S
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ZOLLOTA,VLADIMIR
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/25/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 03:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: LISI

DATE/TIME OF TEST: 01/07/2016 03:25:07

Diagnostics Check		OK	03:20	
Air Blank		0.000	03:20	
Control Test		0.079	03:21	
Air Blank		0.000	03:21	
Subject Sample	# 1		03:23	Subject Test Refused
Breath Volume		0.000	03:23	
Air Blank		0.000	03:23	
Control Test		0.079	03:24	
Air Blank		0.000	03:24	
Diagnostics Check		Fail	03:24	Voltage/Current Test
Air Blank		0.000	03:25	

Data Download Date/Time: 01/30/2016 15:13

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARKOVICH,ROBBIE,S
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ZOLLOTA,VLADIMIR
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/25/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 03:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: LISI

DATE/TIME OF TEST: 01/07/2016 03:31:19

Diagnostics Check		OK	03:28	
Air Blank		0.000	03:28	
Control Test		0.079	03:29	
Air Blank		0.000	03:29	
Subject Sample	# 1		03:29	Subject Test Refused
Breath Volume		0.000	03:29	
Air Blank		0.000	03:30	
Control Test		0.079	03:30	
Air Blank		0.000	03:31	
Diagnostics Check		OK	03:31	

Data Download Date/Time: 01/30/2016 15:13

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARKOVICH,ROBBIE,S
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: WAGNER,WAYNE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/25/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 01:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:24

SUBJECT'S LAST NAME: WATTERS

DATE/TIME OF TEST: 01/08/2016 01:57:52

Diagnostics Check		OK	01:51
Air Blank		0.000	01:51
Control Test		0.079	01:52
Air Blank		0.000	01:52
Subject Sample	# 1	0.179	01:53
Breath Volume		2.773	01:53
Air Blank		0.000	01:53
Air Blank		0.000	01:55
Subject Sample	# 2	0.177	01:56
Breath Volume		1.160	01:56
Air Blank		0.000	01:56
Control Test		0.079	01:57
Air Blank		0.000	01:57
Diagnostics Check		OK	01:57

Data Download Date/Time: 01/30/2016 15:13

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALVEY,DOUGLAS,J
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CARTER,MATTHEW
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/25/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:34

SUBJECT'S LAST NAME: WEEKS

DATE/TIME OF TEST: 01/08/2016 03:06:15

Diagnostics Check		OK	03:02	
Air Blank		0.000	03:02	
Control Test		0.079	03:02	
Air Blank		0.000	03:03	
Subject Sample	# 1		03:04	Subject Test Refused
Breath Volume		0.000	03:04	
Air Blank		0.000	03:05	
Control Test		0.079	03:05	
Air Blank		0.000	03:06	
Diagnostics Check		OK	03:06	

Data Download Date/Time: 01/30/2016 15:13

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARKOVICH,ROBBIE,S
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ZOLLOTA,VLADIMIR
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/25/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 20:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:34

SUBJECT'S LAST NAME: MASINO

DATE/TIME OF TEST: 01/08/2016 21:12:36

Diagnostics Check		OK	21:09	
Air Blank		0.000	21:10	
Control Test		0.079	21:10	
Air Blank		0.000	21:10	
Subject Sample	# 1		21:11	Subject Test Refused
Breath Volume		0.000	21:11	
Air Blank		0.000	21:11	
Control Test		0.079	21:12	
Air Blank		0.000	21:12	
Diagnostics Check		OK	21:12	

Data Download Date/Time: 01/30/2016 15:13

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZOLLOTA,VLADIMIR,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: REEVES,JONATHAN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/25/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 03:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:48

SUBJECT'S LAST NAME: HOWARD

DATE/TIME OF TEST: 01/09/2016 03:19:15

Diagnostics Check	OK	03:17	
Air Blank	0.000	03:18	
Control Test	0.079	03:18	
Air Blank		03:18	RFI Detect
Air Blank	0.000	03:19	

Data Download Date/Time: 01/30/2016 15:13

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZOLLOTA,VLADIMIR,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: REEVES,JONATHAN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/25/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 03:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:48

SUBJECT'S LAST NAME: HOWARD

DATE/TIME OF TEST: 01/09/2016 03:25:15

Diagnostics Check		OK	03:22	
Air Blank		0.000	03:22	
Control Test		0.080	03:22	
Air Blank		0.000	03:23	
Subject Sample	# 1		03:23	Subject Test Refused
Breath Volume		0.000	03:23	
Air Blank		0.000	03:24	
Control Test		0.080	03:24	
Air Blank		0.000	03:25	
Diagnostics Check		OK	03:25	

Data Download Date/Time: 01/30/2016 15:13

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZOLLOTA,VLADIMIR,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ALVEY,DOUGLAS
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/25/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 03:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:04

SUBJECT'S LAST NAME: OFRIA

DATE/TIME OF TEST: 01/09/2016 03:36:57

Diagnostics Check		OK	03:33	
Air Blank		0.000	03:33	
Control Test		0.080	03:34	
Air Blank		0.000	03:34	
Subject Sample	# 1	0.195	03:35	
Breath Volume		3.261	03:35	
Air Blank			03:35	RFI Detect
Air Blank		0.000	03:36	
Control Test		0.080	03:36	
Air Blank		0.000	03:36	
Diagnostics Check		OK	03:36	

Data Download Date/Time: 01/30/2016 15:13

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZOLLOTA,VLADIMIR,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ALVEY,DOUGLAS
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/25/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 03:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:04

SUBJECT'S LAST NAME: OFRIA

DATE/TIME OF TEST: 01/09/2016 03:44:27

Diagnostics Check		OK	03:38
Air Blank		0.000	03:38
Control Test		0.080	03:38
Air Blank		0.000	03:39
Subject Sample	# 1	0.186	03:39
Breath Volume		3.187	03:39
Air Blank		0.000	03:40
Air Blank		0.000	03:42
Subject Sample	# 2	0.180	03:42
Breath Volume		3.230	03:42
Air Blank		0.000	03:43
Control Test		0.077	03:43
Air Blank		0.000	03:44
Diagnostics Check		OK	03:44

Data Download Date/Time: 01/30/2016 15:13

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALVEY,DOUGLAS,J
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ARKOVICH,ROBBIE
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/25/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 23:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:58

SUBJECT'S LAST NAME: LANGSTON-GRAHAM

DATE/TIME OF TEST: 01/15/2016 23:27:29

Diagnostics Check		OK	23:19
Air Blank		0.000	23:20
Control Test		0.079	23:20
Air Blank		0.000	23:21
Subject Sample	# 1	0.211	23:22
Breath Volume		1.871	23:22
Air Blank		0.000	23:22
Air Blank		0.000	23:24
Subject Sample	# 2	0.219	23:25
Breath Volume		2.160	23:25
Air Blank		0.000	23:26
Control Test		0.079	23:26
Air Blank		0.000	23:27
Diagnostics Check		OK	23:27

Data Download Date/Time: 01/30/2016 15:13

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTER,MATTHEW,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ZOLLOTA,VLADIMIR
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/25/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: VILLARREAL

DATE/TIME OF TEST: 01/16/2016 02:40:23

Diagnostics Check		OK	02:34
Air Blank		0.000	02:34
Control Test		0.079	02:34
Air Blank		0.000	02:35
Subject Sample	# 1	0.187	02:35
Breath Volume		1.386	02:35
Air Blank		0.000	02:36
Air Blank		0.000	02:38
Subject Sample	# 2	0.181	02:38
Breath Volume		1.726	02:38
Air Blank		0.000	02:39
Control Test		0.078	02:39
Air Blank		0.000	02:40
Diagnostics Check		OK	02:40

Data Download Date/Time: 01/30/2016 15:13

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROLLESTON,JACOB,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: ROLLESTON,JACOB
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 12/25/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 19:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:27

SUBJECT'S LAST NAME: NGUYEN

DATE/TIME OF TEST: 01/16/2016 20:54:27

Diagnostics Check		OK	20:48
Air Blank		0.000	20:48
Control Test		0.079	20:48
Air Blank		0.000	20:49
Subject Sample	# 1	0.213	20:50
Breath Volume		2.460	20:50
Air Blank		0.000	20:50
Air Blank		0.000	20:52
Subject Sample	# 2	0.205	20:52
Breath Volume		2.250	20:52
Air Blank		0.000	20:53
Control Test		0.079	20:53
Air Blank		0.000	20:54
Diagnostics Check		OK	20:54

Data Download Date/Time: 01/30/2016 15:13

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALVEY,DOUGLAS,J
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ARKOVICH,ROBBIE
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/25/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 13:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:49

SUBJECT'S LAST NAME: FLORENCE

DATE/TIME OF TEST: 01/18/2016 13:36:43

Diagnostics Check		OK	13:33	
Air Blank		0.000	13:34	
Control Test		0.081	13:34	
Air Blank		0.000	13:35	
Subject Sample	# 1		13:35	Subject Test Refused
Breath Volume		0.000	13:35	
Air Blank		0.000	13:35	
Control Test		0.081	13:36	
Air Blank		0.000	13:36	
Diagnostics Check		OK	13:36	

Data Download Date/Time: 01/30/2016 15:13

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 12/14/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:23

SUBJECT'S LAST NAME: MCLEMORE

DATE/TIME OF TEST: 01/01/2016 01:53:27

Diagnostics Check		OK	01:43
Air Blank		0.000	01:44
Control Test		0.080	01:44
Air Blank		0.000	01:45
Subject Sample	# 1	0.183	01:46
Breath Volume		2.339	01:46
Air Blank		0.000	01:46
Air Blank		0.000	01:48
Subject Sample	# 2	0.206	01:48
Breath Volume		2.199	01:48
Air Blank		0.000	01:49
Air Blank		0.000	01:51
Subject Sample	# 3	0.194	01:51
Breath Volume		2.695	01:51
Air Blank		0.000	01:52
Control Test		0.079	01:52
Air Blank		0.000	01:53
Diagnostics Check		OK	01:53

Data Download Date/Time: 01/20/2016 12:06

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 12/14/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: HUTCHINSON

DATE/TIME OF TEST: 01/01/2016 04:47:14

Diagnostics Check		OK	04:40
Air Blank		0.000	04:41
Control Test		0.080	04:41
Air Blank		0.000	04:42
Subject Sample	# 1	0.118	04:42
Breath Volume		2.722	04:42
Air Blank		0.000	04:43
Air Blank		0.000	04:45
Subject Sample	# 2	0.123	04:45
Breath Volume		2.453	04:45
Air Blank		0.000	04:46
Control Test		0.079	04:46
Air Blank		0.000	04:47
Diagnostics Check		OK	04:47

Data Download Date/Time: 01/20/2016 12:06

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THORNSBROUGH,DREW,A
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: JONES,EMMANUEL
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 12/14/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 14:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:46

SUBJECT'S LAST NAME: KELLY

DATE/TIME OF TEST: 01/01/2016 15:13:49

Diagnostics Check		OK	15:07
Air Blank		0.000	15:08
Control Test		0.080	15:08
Air Blank		0.000	15:09
Subject Sample	# 1	0.214	15:09
Breath Volume		1.792	15:09
Air Blank		0.000	15:10
Air Blank		0.000	15:12
Subject Sample	# 2	0.214	15:12
Breath Volume		2.093	15:12
Air Blank		0.000	15:12
Control Test		0.079	15:13
Air Blank		0.000	15:13
Diagnostics Check		OK	15:13

Data Download Date/Time: 01/20/2016 12:06

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 12/14/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:47

SUBJECT'S LAST NAME: BERGFELD

DATE/TIME OF TEST: 01/02/2016 02:19:46

Diagnostics Check		OK	02:16	
Air Blank		0.000	02:16	
Control Test		0.079	02:17	
Air Blank		0.000	02:17	
Subject Sample	# 1		02:18	Subject Test Refused
Breath Volume		0.000	02:18	
Air Blank		0.000	02:18	
Control Test		0.079	02:19	
Air Blank		0.000	02:19	
Diagnostics Check		OK	02:19	

Data Download Date/Time: 01/20/2016 12:06

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICE,JOHN,B
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: RICE,JOHN
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 12/14/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:56

SUBJECT'S LAST NAME: FIEBIG

DATE/TIME OF TEST: 01/03/2016 04:25:38

Diagnostics Check		OK	04:18
Air Blank		0.000	04:19
Control Test		0.080	04:19
Air Blank		0.000	04:20
Subject Sample	# 1	0.121	04:20
Breath Volume		2.687	04:20
Air Blank		0.000	04:21
Air Blank		0.000	04:23
Subject Sample	# 2	0.118	04:24
Breath Volume		1.875	04:24
Air Blank		0.000	04:24
Control Test		0.080	04:25
Air Blank		0.000	04:25
Diagnostics Check		OK	04:25

Data Download Date/Time: 01/20/2016 12:06

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CITRANO,ANTHONY,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: DISNEY,THOMAS
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 12/14/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:32

SUBJECT'S LAST NAME: RAUS

DATE/TIME OF TEST: 01/07/2016 03:59:45

Diagnostics Check		OK	03:53
Air Blank		0.000	03:53
Control Test		0.080	03:54
Air Blank		0.000	03:54
Subject Sample	# 1	0.084	03:55
Breath Volume		1.179	03:55
Air Blank		0.000	03:55
Air Blank		0.000	03:57
Subject Sample	# 2	0.083	03:58
Breath Volume		1.617	03:58
Air Blank		0.000	03:58
Control Test		0.080	03:59
Air Blank		0.000	03:59
Diagnostics Check		OK	03:59

Data Download Date/Time: 01/20/2016 12:06

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 12/14/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:43

SUBJECT'S LAST NAME: LEAKE

DATE/TIME OF TEST: 01/08/2016 04:10:47

Diagnostics Check		OK	04:03
Air Blank		0.000	04:04
Control Test		0.080	04:04
Air Blank		0.000	04:05
Subject Sample	# 1	0.199	04:06
Breath Volume		1.617	04:06
Air Blank		0.000	04:06
Air Blank		0.000	04:08
Subject Sample	# 2	0.183	04:09
Breath Volume		2.066	04:09
Air Blank		0.000	04:09
Control Test		0.079	04:10
Air Blank		0.000	04:10
Diagnostics Check		OK	04:10

Data Download Date/Time: 01/20/2016 12:06

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 12/14/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:41

SUBJECT'S LAST NAME: SILKIN

DATE/TIME OF TEST: 01/08/2016 21:10:27

Diagnostics Check		OK	21:04
Air Blank		0.000	21:04
Control Test		0.080	21:05
Air Blank		0.000	21:05
Subject Sample	# 1	0.000	21:06
Breath Volume		2.765	21:06
Air Blank		0.000	21:06
Air Blank		0.000	21:08
Subject Sample	# 2	0.000	21:09
Breath Volume		2.441	21:09
Air Blank		0.000	21:09
Control Test		0.080	21:09
Air Blank		0.000	21:10
Diagnostics Check		OK	21:10

Data Download Date/Time: 01/20/2016 12:06

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICE,JOHN,B
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: RICE,JOHN
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 12/14/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 19:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:38

SUBJECT'S LAST NAME: FOUCHE

DATE/TIME OF TEST: 01/09/2016 20:05:17

Diagnostics Check		OK	19:58
Air Blank		0.000	19:59
Control Test		0.079	19:59
Air Blank		0.000	20:00
Subject Sample	# 1	0.187	20:00
Breath Volume		2.246	20:00
Air Blank		0.000	20:01
Air Blank		0.000	20:03
Subject Sample	# 2	0.191	20:03
Breath Volume		1.851	20:03
Air Blank		0.000	20:04
Control Test		0.079	20:04
Air Blank		0.000	20:05
Diagnostics Check		OK	20:05

Data Download Date/Time: 01/20/2016 12:06

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 12/14/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:24

SUBJECT'S LAST NAME: BRYANT

DATE/TIME OF TEST: 01/15/2016 03:55:43

Diagnostics Check		OK	03:48
Air Blank		0.000	03:48
Control Test		0.080	03:48
Air Blank		0.000	03:49
Subject Sample	# 1	0.111	03:50
Breath Volume		2.101	03:50
Air Blank		0.000	03:51
Air Blank		0.000	03:53
Subject Sample	# 2	0.114	03:54
Breath Volume		2.109	03:54
Air Blank		0.000	03:54
Control Test		0.080	03:55
Air Blank		0.000	03:55
Diagnostics Check		OK	03:55

Data Download Date/Time: 01/20/2016 12:06

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALAMEDA,JORGE,L
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: 4L26,ALAMEDA
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 12/14/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:25

SUBJECT'S LAST NAME: FELIX GARCIA

DATE/TIME OF TEST: 01/15/2016 05:48:37

Diagnostics Check		OK	05:45	
Air Blank		0.000	05:46	
Control Test		0.080	05:46	
Air Blank		0.000	05:47	
Subject Sample	# 1		05:47	Subject Test Refused
Breath Volume		0.000	05:47	
Air Blank		0.000	05:47	
Control Test		0.080	05:48	
Air Blank		0.000	05:48	
Diagnostics Check		OK	05:48	

Data Download Date/Time: 01/20/2016 12:06

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 12/14/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:09

SUBJECT'S LAST NAME: MANSELL

DATE/TIME OF TEST: 01/17/2016 02:40:39

Diagnostics Check		OK	02:31
Air Blank		0.000	02:32
Control Test		0.080	02:32
Air Blank		0.000	02:33
Subject Sample	# 1	0.155	02:36
Breath Volume		1.750	02:36
Air Blank		0.000	02:36
Air Blank		0.000	02:38
Subject Sample	# 2	0.165	02:39
Breath Volume		1.804	02:39
Air Blank		0.000	02:39
Control Test		0.080	02:40
Air Blank		0.000	02:40
Diagnostics Check		OK	02:40

Data Download Date/Time: 01/20/2016 12:06

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICE,JOHN,B
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: JONES,CHRIS
ARRESTING OFFICER AGENCY: Indian Shores PD
LAST AGENCY INSPECTION: 12/14/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 18:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:20

SUBJECT'S LAST NAME: ASHER

DATE/TIME OF TEST: 01/17/2016 19:48:40

Diagnostics Check		OK	19:41
Air Blank		0.000	19:41
Control Test		0.080	19:41
Air Blank		0.000	19:42
Subject Sample	# 1	0.164	19:43
Breath Volume		1.867	19:43
Air Blank		0.000	19:44
Air Blank		0.000	19:46
Subject Sample	# 2	0.153	19:47
Breath Volume		1.261	19:47
Air Blank		0.000	19:47
Control Test		0.080	19:48
Air Blank		0.000	19:48
Diagnostics Check		OK	19:48

Data Download Date/Time: 01/20/2016 12:06

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICE,JOHN,B
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: RICE,JOHN
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 12/14/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: CLOTWORTHY

DATE/TIME OF TEST: 01/18/2016 21:17:35

Diagnostics Check		OK	21:10
Air Blank		0.000	21:11
Control Test		0.081	21:11
Air Blank		0.000	21:12
Subject Sample	# 1	0.217	21:12
Breath Volume		1.980	21:12
Air Blank		0.000	21:13
Air Blank		0.000	21:15
Subject Sample	# 2	0.217	21:16
Breath Volume		1.835	21:16
Air Blank		0.000	21:16
Control Test		0.080	21:17
Air Blank		0.000	21:17
Diagnostics Check		OK	21:17

Data Download Date/Time: 01/20/2016 12:06

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLICKENSDORF,MICHAEL,R
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: STONER,CYNTHIA
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 01/20/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:58

SUBJECT'S LAST NAME: HOWARD

DATE/TIME OF TEST: 01/20/2016 23:27:04

Diagnostics Check		OK	23:21
Air Blank		0.000	23:21
Control Test		0.081	23:21
Air Blank		0.000	23:22
Subject Sample	# 1	0.208	23:22
Breath Volume		2.464	23:22
Air Blank		0.000	23:23
Air Blank		0.000	23:25
Subject Sample	# 2	0.207	23:25
Breath Volume		2.339	23:25
Air Blank		0.000	23:26
Control Test		0.081	23:26
Air Blank		0.000	23:26
Diagnostics Check		OK	23:27

Data Download Date/Time: 02/22/2016 07:27

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 01/20/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:29

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 01/23/2016 01:58:43

Diagnostics Check		OK	01:54	
Air Blank		0.000	01:55	
Control Test		0.080	01:55	
Air Blank		0.000	01:55	
Subject Sample	# 1		01:57	Subject Test Refused
Breath Volume		0.000	01:57	
Air Blank		0.000	01:57	
Control Test		0.080	01:58	
Air Blank		0.000	01:58	
Diagnostics Check		OK	01:58	

Data Download Date/Time: 02/22/2016 07:27

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 01/20/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:01

SUBJECT'S LAST NAME: OHNA

DATE/TIME OF TEST: 01/23/2016 21:34:47

Diagnostics Check		OK	21:31	
Air Blank		0.000	21:31	
Control Test		0.080	21:31	
Air Blank		0.000	21:32	
Subject Sample	# 1		21:33	Subject Test Refused
Breath Volume		0.000	21:33	
Air Blank		0.000	21:33	
Control Test		0.081	21:34	
Air Blank		0.000	21:34	
Diagnostics Check		OK	21:34	

Data Download Date/Time: 02/22/2016 07:27

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICE,JOHN,B
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: RICE,JOHN
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 01/20/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:12

SUBJECT'S LAST NAME: SANCHEZ

DATE/TIME OF TEST: 01/25/2016 01:40:32

Diagnostics Check		OK	01:33
Air Blank		0.000	01:33
Control Test		0.082	01:34
Air Blank		0.000	01:34
Subject Sample	# 1	0.199	01:35
Breath Volume		1.828	01:35
Air Blank		0.000	01:36
Air Blank		0.000	01:38
Subject Sample	# 2	0.203	01:39
Breath Volume		2.085	01:39
Air Blank		0.000	01:39
Control Test		0.080	01:39
Air Blank		0.000	01:40
Diagnostics Check		OK	01:40

Data Download Date/Time: 02/22/2016 07:27

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICE,JOHN,B
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: RICE,JOHN
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 01/20/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:44

SUBJECT'S LAST NAME: ISKRA

DATE/TIME OF TEST: 01/26/2016 02:24:52

Diagnostics Check		OK	02:14	
Air Blank		0.000	02:14	
Control Test		0.080	02:15	
Air Blank		0.000	02:15	
Subject Sample	# 1		02:18	No Sample Provided
Breath Volume		0.000	02:18	
Air Blank		0.000	02:19	
Air Blank		0.000	02:21	
Subject Sample	# 2		02:23	Subject Test Refused
Breath Volume		0.000	02:23	
Air Blank		0.000	02:23	
Control Test		0.080	02:24	
Air Blank		0.000	02:24	
Diagnostics Check		OK	02:24	

Data Download Date/Time: 02/22/2016 07:27

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CALDERONE,STEVEN,A
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: CALDERONE,STEVEN
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 01/20/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 06:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:56

SUBJECT'S LAST NAME: WHALEN

DATE/TIME OF TEST: 01/26/2016 07:26:49

Diagnostics Check		OK	07:19
Air Blank		0.000	07:20
Control Test		0.080	07:20
Air Blank		0.000	07:21
Subject Sample	# 1	0.229	07:21
Breath Volume		1.429	07:21
Air Blank		0.000	07:22
Air Blank		0.000	07:24
Subject Sample	# 2	0.235	07:25
Breath Volume		1.261	07:25
Air Blank		0.000	07:25
Control Test		0.081	07:26
Air Blank		0.000	07:26
Diagnostics Check		OK	07:26

Data Download Date/Time: 02/22/2016 07:27

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 01/20/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: KELLY

DATE/TIME OF TEST: 01/31/2016 01:07:26

Diagnostics Check		OK	01:01
Air Blank		0.000	01:01
Control Test		0.080	01:01
Air Blank		0.000	01:02
Subject Sample	# 1	0.178	01:03
Breath Volume		2.984	01:03
Air Blank		0.000	01:03
Air Blank		0.000	01:05
Subject Sample	# 2	0.171	01:05
Breath Volume		1.980	01:05
Air Blank		0.000	01:06
Control Test		0.080	01:06
Air Blank		0.000	01:07
Diagnostics Check		OK	01:07

Data Download Date/Time: 02/22/2016 07:27

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICE,JOHN,B
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: RICE,JOHN
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 01/20/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:57

SUBJECT'S LAST NAME: BEATTY

DATE/TIME OF TEST: 01/31/2016 02:27:46

Diagnostics Check		OK	02:20
Air Blank		0.000	02:20
Control Test		0.080	02:21
Air Blank		0.000	02:21
Subject Sample	# 1	0.203	02:22
Breath Volume		2.839	02:22
Air Blank		0.000	02:22
Air Blank		0.000	02:24
Subject Sample	# 2	0.202	02:26
Breath Volume		1.945	02:26
Air Blank		0.000	02:26
Control Test		0.080	02:27
Air Blank		0.000	02:27
Diagnostics Check		OK	02:27

Data Download Date/Time: 02/22/2016 07:27

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 01/20/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:25

SUBJECT'S LAST NAME: VALDES

DATE/TIME OF TEST: 01/31/2016 05:53:07

Diagnostics Check		OK	05:45
Air Blank		0.000	05:46
Control Test		0.080	05:46
Air Blank		0.000	05:47
Subject Sample	# 1	0.189	05:48
Breath Volume		1.796	05:48
Air Blank		0.000	05:48
Air Blank		0.000	05:50
Subject Sample	# 2	0.184	05:51
Breath Volume		1.683	05:51
Air Blank		0.000	05:52
Control Test		0.080	05:52
Air Blank		0.000	05:53
Diagnostics Check		OK	05:53

Data Download Date/Time: 02/22/2016 07:27

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KHAN,WAKEEL,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: KHAN,WAKEEL
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/02/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 05:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:08

SUBJECT'S LAST NAME: WATSON

DATE/TIME OF TEST: 01/01/2016 06:47:24

Diagnostics Check		OK	06:42	
Air Blank		0.000	06:42	
Control Test		0.079	06:43	
Air Blank		0.000	06:43	
Subject Sample	# 1		06:45	RFI Detect
Breath Volume		0.000	06:45	
Air Blank		0.000	06:46	
Control Test		0.079	06:46	
Air Blank			06:46	RFI Detect
Air Blank		0.000	06:47	

Data Download Date/Time: 01/05/2016 12:09

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PHILLIPS,JORDAN,W
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: JONES,HUGH
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/02/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 19:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:14

SUBJECT'S LAST NAME: WOODALL

DATE/TIME OF TEST: 01/02/2016 20:48:51

Diagnostics Check		OK	20:45	
Air Blank		0.000	20:45	
Control Test		0.079	20:45	
Air Blank		0.000	20:46	
Subject Sample	# 1		20:47	Subject Test Refused
Breath Volume		0.058	20:47	
Air Blank		0.000	20:47	
Control Test		0.079	20:48	
Air Blank		0.000	20:48	
Diagnostics Check		OK	20:48	

Data Download Date/Time: 01/05/2016 12:09

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PHILLIPS,JORDAN,W
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: MUNOZ,JON
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/02/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: HEFLIN

DATE/TIME OF TEST: 01/02/2016 22:32:34

Diagnostics Check		OK	22:25
Air Blank		0.000	22:26
Control Test		0.078	22:26
Air Blank		0.000	22:27
Subject Sample	# 1	0.223	22:27
Breath Volume		1.867	22:27
Air Blank		0.000	22:28
Air Blank		0.000	22:30
Subject Sample	# 2	0.228	22:30
Breath Volume		2.703	22:30
Air Blank		0.000	22:31
Control Test		0.078	22:31
Air Blank		0.000	22:32
Diagnostics Check		OK	22:32

Data Download Date/Time: 01/05/2016 12:09

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,JOSEPH,W
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: BAEZ,NELLY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/02/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: VAVRA

DATE/TIME OF TEST: 01/04/2016 22:28:05

Diagnostics Check		OK	22:21
Air Blank		0.000	22:21
Control Test		0.078	22:22
Air Blank		0.000	22:22
Subject Sample	# 1	0.160	22:23
Breath Volume		1.292	22:23
Air Blank		0.000	22:23
Air Blank		0.000	22:25
Subject Sample	# 2	0.158	22:26
Breath Volume		1.335	22:26
Air Blank		0.000	22:27
Control Test		0.077	22:27
Air Blank		0.000	22:27
Diagnostics Check		OK	22:28

Data Download Date/Time: 01/05/2016 12:09

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUCAS,NATHANIEL,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: LUCAS,NATHANIEL
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/02/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 01/05/2016 01:40:12

Diagnostics Check		OK	01:33
Air Blank		0.000	01:33
Control Test		0.079	01:33
Air Blank		0.000	01:34
Subject Sample	# 1	0.125	01:35
Breath Volume		1.710	01:35
Air Blank		0.000	01:36
Air Blank		0.000	01:37
Subject Sample	# 2	0.121	01:38
Breath Volume		1.652	01:38
Air Blank		0.000	01:39
Control Test		0.079	01:39
Air Blank		0.000	01:40
Diagnostics Check		OK	01:40

Data Download Date/Time: 01/05/2016 12:09

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MELTZER,CHRISTOPHER,J
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: MELTZER,CHRISTOPHER
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 01/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: STEVERSON

DATE/TIME OF TEST: 01/06/2016 04:07:39

Diagnostics Check		OK	04:00
Air Blank		0.000	04:01
Control Test		0.078	04:01
Air Blank		0.000	04:01
Subject Sample	# 1	0.190	04:02
Breath Volume		1.937	04:02
Air Blank		0.000	04:03
Air Blank		0.000	04:05
Subject Sample	# 2	0.182	04:06
Breath Volume		2.425	04:06
Air Blank		0.000	04:06
Control Test		0.077	04:07
Air Blank		0.000	04:07
Diagnostics Check		OK	04:07

Data Download Date/Time: 02/05/2016 11:45

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PHILLIPS,JORDAN,W
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: BROOME,DEAONTE
ARRESTING OFFICER AGENCY: Bartow PD
LAST AGENCY INSPECTION: 01/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:38

SUBJECT'S LAST NAME: PRZEKOP

DATE/TIME OF TEST: 01/07/2016 23:13:43

Diagnostics Check		OK	23:02
Air Blank		0.000	23:02
Control Test		0.079	23:02
Air Blank		0.000	23:03
Subject Sample	# 1	0.125	23:06
Breath Volume		1.398	23:06
Air Blank		0.000	23:07
Air Blank		0.000	23:08
Subject Sample	# 2	0.120	23:12
Breath Volume		1.203	23:12
Air Blank		0.000	23:12
Control Test		0.078	23:13
Air Blank		0.000	23:13
Diagnostics Check		OK	23:13

Data Download Date/Time: 02/05/2016 11:45

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUCAS,NATHANIEL,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: LUCAS,NATHANIEL
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 01/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:10

SUBJECT'S LAST NAME: BANDA

DATE/TIME OF TEST: 01/08/2016 21:49:36

Diagnostics Check		OK	21:43
Air Blank		0.000	21:43
Control Test		0.078	21:44
Air Blank		0.000	21:44
Subject Sample	# 1	0.170	21:45
Breath Volume		1.597	21:45
Air Blank		0.000	21:45
Air Blank		0.000	21:47
Subject Sample	# 2	0.156	21:48
Breath Volume		1.796	21:48
Air Blank		0.000	21:48
Control Test		0.077	21:49
Air Blank		0.000	21:49
Diagnostics Check		OK	21:49

Data Download Date/Time: 02/05/2016 11:45

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUCAS,NATHANIEL,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: LUCAS,NATHANIEL
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 01/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: CRAWFORD

DATE/TIME OF TEST: 01/14/2016 00:55:49

Diagnostics Check	OK	00:54	
Air Blank	0.000	00:54	
Control Test	0.078	00:55	
Air Blank		00:55	RFI Detect
Air Blank	0.000	00:55	

Data Download Date/Time: 02/05/2016 11:45

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUCAS,NATHANIEL,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: LUCAS,NATHANIEL
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 01/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: CRAWFORD

DATE/TIME OF TEST: 01/14/2016 01:06:02

Diagnostics Check		OK	00:59
Air Blank		0.000	00:59
Control Test		0.078	00:59
Air Blank		0.000	01:00
Subject Sample	# 1	0.217	01:01
Breath Volume		1.449	01:01
Air Blank		0.000	01:02
Air Blank		0.000	01:03
Subject Sample	# 2	0.215	01:04
Breath Volume		1.312	01:04
Air Blank		0.000	01:05
Control Test		0.077	01:05
Air Blank		0.000	01:05
Diagnostics Check		OK	01:06

Data Download Date/Time: 02/05/2016 11:45

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JUDAH,PATTY,A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: HEWETT,MIKE
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 01/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: YANNONE

DATE/TIME OF TEST: 01/15/2016 03:40:04

Diagnostics Check		OK	03:37	
Air Blank		0.000	03:38	
Control Test		0.079	03:38	
Air Blank		0.000	03:39	
Subject Sample	# 1		03:39	Subject Test Refused
Breath Volume		0.000	03:39	
Air Blank			03:39	RFI Detect
Air Blank		0.000	03:40	

Data Download Date/Time: 02/05/2016 11:45

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JUDAH,PATTY,A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: HEWETT,MIKE
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 01/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: YANNONE

DATE/TIME OF TEST: 01/15/2016 03:46:00

Diagnostics Check		OK	03:43	
Air Blank		0.000	03:43	
Control Test		0.079	03:43	
Air Blank		0.000	03:44	
Subject Sample	# 1		03:44	Subject Test Refused
Breath Volume		0.000	03:44	
Air Blank		0.000	03:45	
Control Test		0.079	03:45	
Air Blank		0.000	03:45	
Diagnostics Check		OK	03:46	

Data Download Date/Time: 02/05/2016 11:45

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TOKAR,ROBERT,J
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: TOKAR,R
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 01/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 12:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:05

SUBJECT'S LAST NAME: QUIRK

DATE/TIME OF TEST: 01/16/2016 13:42:46

Diagnostics Check		OK	13:36
Air Blank		0.000	13:37
Control Test		0.078	13:37
Air Blank		0.000	13:38
Subject Sample	# 1	0.000	13:38
Breath Volume		2.332	13:38
Air Blank		0.000	13:39
Air Blank		0.000	13:41
Subject Sample	# 2	0.000	13:41
Breath Volume		1.621	13:41
Air Blank		0.000	13:41
Control Test		0.076	13:42
Air Blank		0.000	13:42
Diagnostics Check		OK	13:42

Data Download Date/Time: 02/05/2016 11:45

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LACAPRA,DANIEL,C
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: LACAPRA,DANIEL
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 01/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 11:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:00

SUBJECT'S LAST NAME: QUIRK

DATE/TIME OF TEST: 01/18/2016 12:30:54

Diagnostics Check		OK	12:25
Air Blank		0.000	12:25
Control Test		0.079	12:25
Air Blank		0.000	12:26
Subject Sample	# 1	0.000	12:26
Breath Volume		1.824	12:26
Air Blank		0.000	12:27
Air Blank		0.000	12:29
Subject Sample	# 2	0.000	12:29
Breath Volume		1.597	12:29
Air Blank		0.000	12:29
Control Test		0.077	12:30
Air Blank		0.000	12:30
Diagnostics Check		OK	12:30

Data Download Date/Time: 02/05/2016 11:45

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WIEGERT,KIMBRA,L
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: SERRATO,CESAR
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 01/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 09:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:50

SUBJECT'S LAST NAME: SYLVESTER

DATE/TIME OF TEST: 01/19/2016 11:24:24

Diagnostics Check		OK	11:21	
Air Blank		0.000	11:22	
Control Test		0.080	11:22	
Air Blank		0.000	11:22	
Subject Sample	# 1		11:23	Subject Test Refused
Breath Volume		0.000	11:23	
Air Blank		0.000	11:23	
Control Test		0.079	11:23	
Air Blank		0.000	11:24	
Diagnostics Check		OK	11:24	

Data Download Date/Time: 02/05/2016 11:45

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JUDAH,PATTY,A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: HEWETT,MIKE
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 01/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: BARKER

DATE/TIME OF TEST: 01/20/2016 02:13:37

Diagnostics Check		OK	02:01	
Air Blank		0.000	02:02	
Control Test		0.080	02:02	
Air Blank		0.000	02:03	
Subject Sample	# 1	0.219	02:06	Volume Not Met
Breath Volume		0.035	02:06	
Air Blank		0.000	02:07	
Air Blank		0.000	02:08	
Subject Sample	# 2	0.191	02:11	Volume Not Met
Breath Volume		0.570	02:11	
Air Blank		0.000	02:12	
Control Test		0.079	02:12	
Air Blank		0.000	02:13	
Diagnostics Check		OK	02:13	

Data Download Date/Time: 02/05/2016 11:45

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELANEY,AMY,L
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: TURNER,L
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 01/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 05:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:09

SUBJECT'S LAST NAME: THREATT

DATE/TIME OF TEST: 01/20/2016 07:40:27

Diagnostics Check	OK	07:39	
Air Blank		07:39	RFI Detect
Air Blank	0.000	07:40	

Data Download Date/Time: 02/05/2016 11:45

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELANEY,AMY,L
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: TURNER,L
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 01/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 05:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:09

SUBJECT'S LAST NAME: THREATT

DATE/TIME OF TEST: 01/20/2016 07:48:59

Diagnostics Check		OK	07:42
Air Blank		0.000	07:42
Control Test		0.079	07:42
Air Blank		0.000	07:43
Subject Sample	# 1	0.000	07:44
Breath Volume		1.445	07:44
Air Blank		0.000	07:44
Air Blank		0.000	07:46
Subject Sample	# 2	0.000	07:47
Breath Volume		1.589	07:47
Air Blank		0.000	07:48
Control Test		0.080	07:48
Air Blank		0.000	07:48
Diagnostics Check		OK	07:48

Data Download Date/Time: 02/05/2016 11:45

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHURCH,JOHN,M
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: CHURCH,JOHN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 01/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: CLEMENTS

DATE/TIME OF TEST: 01/23/2016 23:34:57

Diagnostics Check		OK	23:32	
Air Blank		0.000	23:32	
Control Test		0.080	23:32	
Air Blank		0.000	23:33	
Subject Sample	# 1		23:33	Subject Test Refused
Breath Volume		0.000	23:33	
Air Blank		0.000	23:34	
Control Test		0.079	23:34	
Air Blank		0.000	23:34	
Diagnostics Check		OK	23:34	

Data Download Date/Time: 02/05/2016 11:45

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUCAS,NATHANIEL,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: LUCAS,NATHANIEL
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 01/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: VILLA

DATE/TIME OF TEST: 01/24/2016 02:58:53

Diagnostics Check		OK	02:44	
Air Blank		0.000	02:44	
Control Test		0.080	02:44	
Air Blank		0.000	02:45	
Subject Sample	# 1	0.206	02:48	
Breath Volume		1.101	02:48	
Air Blank		0.000	02:48	
Air Blank		0.000	02:50	
Subject Sample	# 2	0.157	02:53	Volume Not Met
Breath Volume		1.058	02:53	
Air Blank		0.000	02:54	
Air Blank		0.000	02:56	
Subject Sample	# 3		02:57	Subject Test Refused
Breath Volume		0.320	02:57	
Air Blank		0.000	02:57	
Control Test		0.075	02:58	
Air Blank		0.000	02:58	
Diagnostics Check		OK	02:58	

Data Download Date/Time: 02/05/2016 11:45

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHURCH,JOHN,M
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: CHURCH,JOHN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 01/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:28

SUBJECT'S LAST NAME: BAILEY

DATE/TIME OF TEST: 01/24/2016 21:54:30

Diagnostics Check		OK	21:51	
Air Blank		0.000	21:52	
Control Test		0.080	21:52	
Air Blank		0.000	21:52	
Subject Sample	# 1		21:53	Subject Test Refused
Breath Volume		0.000	21:53	
Air Blank		0.000	21:53	
Control Test		0.079	21:53	
Air Blank		0.000	21:54	
Diagnostics Check		OK	21:54	

Data Download Date/Time: 02/05/2016 11:45

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHURCH,JOHN,M
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: CHURCH,JOHN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 01/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:54

SUBJECT'S LAST NAME: GRAY

DATE/TIME OF TEST: 01/25/2016 02:25:16

Diagnostics Check		OK	02:22	
Air Blank		0.000	02:22	
Control Test		0.079	02:23	
Air Blank		0.000	02:23	
Subject Sample	# 1		02:23	Subject Test Refused
Breath Volume		0.000	02:23	
Air Blank		0.000	02:24	
Control Test		0.079	02:24	
Air Blank		0.000	02:25	
Diagnostics Check		OK	02:25	

Data Download Date/Time: 02/05/2016 11:45

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PHILLIPS,JORDAN,W
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: CANNON,NOAH
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 01/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:57

SUBJECT'S LAST NAME: MOORE

DATE/TIME OF TEST: 01/25/2016 22:36:24

Diagnostics Check		OK	22:28
Air Blank		0.000	22:28
Control Test		0.078	22:28
Air Blank		0.000	22:29
Subject Sample	# 1	0.119	22:31
Breath Volume		1.777	22:31
Air Blank		0.000	22:32
Air Blank		0.000	22:34
Subject Sample	# 2	0.114	22:34
Breath Volume		2.203	22:34
Air Blank		0.000	22:35
Control Test		0.076	22:35
Air Blank		0.000	22:36
Diagnostics Check		OK	22:36

Data Download Date/Time: 02/05/2016 11:45

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARREN,BILLY,J
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: WARREN,BILLY
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 01/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:47

SUBJECT'S LAST NAME: HIRSCH

DATE/TIME OF TEST: 01/26/2016 23:57:01

Diagnostics Check		OK	23:45	
Air Blank		0.000	23:46	
Control Test		0.079	23:46	
Air Blank		0.000	23:46	
Subject Sample	# 1		23:50	No Sample Provided
Breath Volume		0.000	23:50	
Air Blank		0.000	23:50	
Air Blank		0.000	23:52	
Subject Sample	# 2		23:55	No Sample Provided
Breath Volume		0.000	23:55	
Air Blank		0.000	23:56	
Control Test		0.079	23:56	
Air Blank		0.000	23:56	
Diagnostics Check		OK	23:57	

Data Download Date/Time: 02/05/2016 11:45

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUCAS,NATHANIEL,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: LUCAS,NATHANIEL
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 01/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: PELZL

DATE/TIME OF TEST: 01/28/2016 02:23:13

Diagnostics Check		OK	02:17	
Air Blank		0.000	02:17	
Control Test		0.078	02:18	
Air Blank		0.000	02:18	
Subject Sample	# 1	0.000	02:18	
Breath Volume		1.492	02:18	
Air Blank		0.000	02:19	
Air Blank		0.000	02:21	
Subject Sample	# 2		02:21	RFI Detect
Breath Volume		1.687	02:21	
Air Blank		0.000	02:22	
Control Test		0.080	02:22	
Air Blank		0.000	02:23	
Diagnostics Check		OK	02:23	

Data Download Date/Time: 02/05/2016 11:45

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUCAS,NATHANIEL,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: LUCAS,NATHANIEL
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 01/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: PELZL

DATE/TIME OF TEST: 01/28/2016 02:31:10

Diagnostics Check		OK	02:25
Air Blank		0.000	02:25
Control Test		0.080	02:26
Air Blank		0.000	02:26
Subject Sample	# 1	0.000	02:26
Breath Volume		1.558	02:26
Air Blank		0.000	02:27
Air Blank		0.000	02:29
Subject Sample	# 2	0.000	02:29
Breath Volume		1.105	02:29
Air Blank		0.000	02:30
Control Test		0.078	02:30
Air Blank		0.000	02:31
Diagnostics Check		OK	02:31

Data Download Date/Time: 02/05/2016 11:45

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUCAS,NATHANIEL,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: CHURCH,JOHN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 01/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: HOOD

DATE/TIME OF TEST: 01/28/2016 06:09:07

Diagnostics Check	OK	06:07	
Air Blank	0.000	06:08	
Control Test	0.092	06:08	Control Outside Tolerance
Air Blank	0.000	06:09	

Data Download Date/Time: 02/05/2016 11:45

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUCAS,NATHANIEL,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: CHURCH,JOHN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 01/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: HOOD

DATE/TIME OF TEST: 01/28/2016 06:14:28

Diagnostics Check		OK	06:11	
Air Blank		0.000	06:12	
Control Test		0.078	06:12	
Air Blank		0.000	06:12	
Subject Sample	# 1		06:13	Subject Test Refused
Breath Volume		0.000	06:13	
Air Blank		0.000	06:13	
Control Test		0.078	06:13	
Air Blank		0.000	06:14	
Diagnostics Check		OK	06:14	

Data Download Date/Time: 02/05/2016 11:45

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARRIER,JAMES,D
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: CARRIER,JAMES
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 01/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:03

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 01/30/2016 21:44:00

Diagnostics Check		OK	21:32	
Air Blank		0.000	21:33	
Control Test		0.077	21:33	
Air Blank		0.000	21:33	
Subject Sample	# 1		21:37	No Sample Provided
Breath Volume		0.000	21:37	
Air Blank		0.000	21:37	
Air Blank		0.000	21:39	
Subject Sample	# 2		21:42	No Sample Provided
Breath Volume		0.000	21:42	
Air Blank		0.000	21:43	
Control Test		0.079	21:43	
Air Blank		0.000	21:43	
Diagnostics Check		OK	21:44	

Data Download Date/Time: 02/05/2016 11:45

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PHILLIPS,JORDAN,W
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: FRANCE,PAUL
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: PILIARIS

DATE/TIME OF TEST: 01/31/2016 01:07:18

Diagnostics Check		OK	01:01
Air Blank		0.000	01:01
Control Test		0.079	01:01
Air Blank		0.000	01:02
Subject Sample	# 1	0.000	01:02
Breath Volume		2.394	01:02
Air Blank		0.000	01:03
Air Blank		0.000	01:05
Subject Sample	# 2	0.000	01:05
Breath Volume		2.457	01:05
Air Blank		0.000	01:06
Control Test		0.079	01:06
Air Blank		0.000	01:07
Diagnostics Check		OK	01:07

Data Download Date/Time: 02/05/2016 11:45

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIVERA,ANGELO,L
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: RIVERA,ANGELO
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 01/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:21

SUBJECT'S LAST NAME: BOATNER

DATE/TIME OF TEST: 01/31/2016 22:56:39

Diagnostics Check		OK	22:48
Air Blank		0.000	22:49
Control Test		0.078	22:49
Air Blank		0.000	22:49
Subject Sample	# 1	0.213	22:51
Breath Volume		1.761	22:51
Air Blank		0.000	22:51
Air Blank		0.000	22:53
Subject Sample	# 2	0.220	22:55
Breath Volume		1.625	22:55
Air Blank		0.000	22:55
Control Test		0.078	22:56
Air Blank		0.000	22:56
Diagnostics Check		OK	22:56

Data Download Date/Time: 02/05/2016 11:45

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRIOLAT,CASEY,J
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: BRIOLAT,CASEY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/02/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A1
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: LORENZ

DATE/TIME OF TEST: 01/03/2016 00:53:55

Diagnostics Check		OK	00:51	
Air Blank		0.000	00:51	
Control Test		0.079	00:51	
Air Blank		0.000	00:52	
Subject Sample	# 1		00:52	Subject Test Refused
Breath Volume		0.000	00:52	
Air Blank		0.000	00:53	
Control Test		0.078	00:53	
Air Blank		0.000	00:53	
Diagnostics Check		OK	00:53	

Data Download Date/Time: 01/19/2016 13:35

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONSEGUR,ROBERT,
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: NIXON,BARRY
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 12/02/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A1
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:44

SUBJECT'S LAST NAME: BRAHM

DATE/TIME OF TEST: 01/07/2016 00:10:54

Diagnostics Check		OK	00:04
Air Blank		0.000	00:05
Control Test		0.079	00:05
Air Blank		0.000	00:06
Subject Sample	# 1	0.000	00:06
Breath Volume		3.121	00:06
Air Blank		0.000	00:07
Air Blank		0.000	00:09
Subject Sample	# 2	0.000	00:09
Breath Volume		2.183	00:09
Air Blank		0.000	00:10
Control Test		0.079	00:10
Air Blank		0.000	00:10
Diagnostics Check		OK	00:10

Data Download Date/Time: 01/19/2016 13:35

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSENBLOOM,MATTHEW,S
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: ROSENBLOOM,MATTHEW
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 12/02/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A1
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:19

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 01/09/2016 03:43:20

Diagnostics Check		OK	03:40	
Air Blank		0.000	03:40	
Control Test		0.078	03:40	
Air Blank		0.000	03:41	
Subject Sample	# 1		03:41	RFI Detect
Breath Volume		0.000	03:41	
Air Blank		0.000	03:42	
Control Test		0.079	03:42	
Air Blank		0.000	03:43	
Diagnostics Check		OK	03:43	

Data Download Date/Time: 01/19/2016 13:35

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSENBLOOM,MATTHEW,S
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: ROSENBLOOM,MATTHEW
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 12/02/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A1
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:19

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 01/09/2016 03:51:44

Diagnostics Check		OK	03:44
Air Blank		0.000	03:45
Control Test		0.079	03:45
Air Blank		0.000	03:46
Subject Sample	# 1	0.176	03:46
Breath Volume		2.824	03:46
Air Blank		0.000	03:47
Air Blank		0.000	03:49
Subject Sample	# 2	0.189	03:50
Breath Volume		3.871	03:50
Air Blank		0.000	03:50
Control Test		0.079	03:51
Air Blank		0.000	03:51
Diagnostics Check		OK	03:51

Data Download Date/Time: 01/19/2016 13:35

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENBO,A,R
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: NIXON,B
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 12/02/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A1
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: WILLIAMSON

DATE/TIME OF TEST: 01/16/2016 04:02:26

Diagnostics Check		OK	03:56
Air Blank		0.000	03:56
Control Test		0.077	03:56
Air Blank		0.000	03:57
Subject Sample	# 1	0.170	03:57
Breath Volume		2.414	03:57
Air Blank		0.000	03:58
Air Blank		0.000	04:00
Subject Sample	# 2	0.174	04:00
Breath Volume		3.101	04:00
Air Blank		0.000	04:01
Control Test		0.078	04:01
Air Blank		0.000	04:02
Diagnostics Check		OK	04:02

Data Download Date/Time: 01/19/2016 13:35

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALVARADO,ANGEL,L
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: ALVARADO,ANGEL
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/19/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A1
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 06:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:20

SUBJECT'S LAST NAME: ZENQUIS

DATE/TIME OF TEST: 01/21/2016 07:49:35

Diagnostics Check		OK	07:43
Air Blank		0.000	07:44
Control Test		0.078	07:44
Air Blank		0.000	07:44
Subject Sample	# 1	0.084	07:45
Breath Volume		2.085	07:45
Air Blank		0.000	07:45
Air Blank		0.000	07:47
Subject Sample	# 2	0.087	07:48
Breath Volume		2.152	07:48
Air Blank		0.000	07:48
Control Test		0.080	07:49
Air Blank		0.000	07:49
Diagnostics Check		OK	07:49

Data Download Date/Time: 02/25/2016 11:57

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001082

INSTRUMENT REGISTERED TO: PLANTATION PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEELE,JACOB,C
BREATH TEST OPERATOR AGENCY: Plantation PD
ARRESTING OFFICER: POWELL,CHRISTOPHER
ARRESTING OFFICER AGENCY: Plantation PD
LAST AGENCY INSPECTION: 12/26/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26614080A7
DRY GAS EXPIRY DATE: 10/05/2016
VIOLATION CODE: DUI
ARREST TIME: 06:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:36

SUBJECT'S LAST NAME: HARRIOTT

DATE/TIME OF TEST: 01/06/2016 06:59:21

Diagnostics Check	OK	06:58	
Air Blank		06:58	Ambient Fail
Air Blank	0.000	06:59	

Data Download Date/Time: 01/27/2016 13:42

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001082

INSTRUMENT REGISTERED TO: PLANTATION PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEELE,JACOB,C
BREATH TEST OPERATOR AGENCY: Plantation PD
ARRESTING OFFICER: POWELL,CHRISTOPHER
ARRESTING OFFICER AGENCY: Plantation PD
LAST AGENCY INSPECTION: 12/26/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26614080A7
DRY GAS EXPIRY DATE: 10/05/2016
VIOLATION CODE: DUI
ARREST TIME: 06:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:36

SUBJECT'S LAST NAME: HARRIOTT

DATE/TIME OF TEST: 01/06/2016 07:09:56

Diagnostics Check		OK	07:02
Air Blank		0.000	07:03
Control Test		0.078	07:03
Air Blank		0.000	07:04
Subject Sample	# 1	0.228	07:05
Breath Volume		2.230	07:05
Air Blank		0.000	07:05
Air Blank		0.000	07:07
Subject Sample	# 2	0.225	07:08
Breath Volume		2.203	07:08
Air Blank		0.000	07:09
Control Test		0.077	07:09
Air Blank		0.000	07:09
Diagnostics Check		OK	07:09

Data Download Date/Time: 01/27/2016 13:42

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001082

INSTRUMENT REGISTERED TO: PLANTATION PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PADRON JR,RICARDO,
BREATH TEST OPERATOR AGENCY: Plantation PD
ARRESTING OFFICER: SHEPPARD,BRITTANY
ARRESTING OFFICER AGENCY: Plantation PD
LAST AGENCY INSPECTION: 12/26/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26614080A7
DRY GAS EXPIRY DATE: 10/05/2016
VIOLATION CODE: DUI
ARREST TIME: 23:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:03

SUBJECT'S LAST NAME: MORALES

DATE/TIME OF TEST: 01/17/2016 01:29:58

Diagnostics Check		OK	01:23
Air Blank		0.000	01:24
Control Test		0.077	01:24
Air Blank		0.000	01:25
Subject Sample	# 1	0.099	01:25
Breath Volume		1.835	01:25
Air Blank		0.000	01:26
Air Blank		0.000	01:28
Subject Sample	# 2	0.096	01:28
Breath Volume		1.750	01:28
Air Blank		0.000	01:29
Control Test		0.077	01:29
Air Blank		0.000	01:29
Diagnostics Check		OK	01:29

Data Download Date/Time: 01/27/2016 13:42

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001082

INSTRUMENT REGISTERED TO: PLANTATION PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAETERN,MEUY,
BREATH TEST OPERATOR AGENCY: Plantation PD
ARRESTING OFFICER: POWELL,CHRISTOPHER
ARRESTING OFFICER AGENCY: Plantation PD
LAST AGENCY INSPECTION: 12/26/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26614080A7
DRY GAS EXPIRY DATE: 10/05/2016
VIOLATION CODE: DUI
ARREST TIME: 03:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:51

SUBJECT'S LAST NAME: WHILBY

DATE/TIME OF TEST: 01/26/2016 04:20:54

Diagnostics Check		OK	04:11
Air Blank		0.000	04:12
Control Test		0.078	04:12
Air Blank		0.000	04:13
Subject Sample	# 1	0.186	04:15
Breath Volume		1.304	04:15
Air Blank		0.000	04:15
Air Blank		0.000	04:17
Subject Sample	# 2	0.185	04:19
Breath Volume		1.492	04:19
Air Blank		0.000	04:19
Control Test		0.076	04:20
Air Blank		0.000	04:20
Diagnostics Check		OK	04:20

Data Download Date/Time: 01/27/2016 13:42

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001082

INSTRUMENT REGISTERED TO: PLANTATION PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PADRON,RICARDO,
BREATH TEST OPERATOR AGENCY: Plantation PD
ARRESTING OFFICER: MULLINGS,BASIL
ARRESTING OFFICER AGENCY: Plantation PD
LAST AGENCY INSPECTION: 01/27/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26614080A7
DRY GAS EXPIRY DATE: 10/05/2016
VIOLATION CODE: DUI
ARREST TIME: 03:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:06

SUBJECT'S LAST NAME: FANN

DATE/TIME OF TEST: 01/31/2016 04:30:06

Diagnostics Check		OK	04:27	
Air Blank		0.000	04:27	
Control Test		0.077	04:28	
Air Blank		0.000	04:28	
Subject Sample	# 1		04:28	Subject Test Refused
Breath Volume		0.000	04:28	
Air Blank		0.000	04:29	
Control Test		0.077	04:29	
Air Blank		0.000	04:30	
Diagnostics Check		OK	04:30	

Data Download Date/Time: 02/26/2016 12:55

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEVERENZ,PATRICK,T
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: TERRELL,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:58

SUBJECT'S LAST NAME: TOWNES

DATE/TIME OF TEST: 01/01/2016 00:25:05

Diagnostics Check		OK	00:18
Air Blank		0.000	00:19
Control Test		0.078	00:19
Air Blank		0.000	00:20
Subject Sample	# 1	0.168	00:20
Breath Volume		1.878	00:20
Air Blank		0.000	00:21
Air Blank		0.000	00:23
Subject Sample	# 2	0.166	00:23
Breath Volume		1.968	00:23
Air Blank		0.000	00:24
Control Test		0.079	00:24
Air Blank		0.000	00:25
Diagnostics Check		OK	00:25

Data Download Date/Time: 01/15/2016 09:08

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEVERENZ,PATRICK,T
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HOOVER,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:41

SUBJECT'S LAST NAME: RIEMER

DATE/TIME OF TEST: 01/01/2016 03:08:06

Diagnostics Check		OK	03:01
Air Blank		0.000	03:02
Control Test		0.079	03:02
Air Blank		0.000	03:03
Subject Sample	# 1	0.125	03:03
Breath Volume		2.437	03:03
Air Blank		0.000	03:04
Air Blank		0.000	03:06
Subject Sample	# 2	0.135	03:06
Breath Volume		1.656	03:06
Air Blank		0.000	03:07
Control Test		0.079	03:07
Air Blank		0.000	03:08
Diagnostics Check		OK	03:08

Data Download Date/Time: 01/15/2016 09:08

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEVERENZ,PATRICK,T
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MONTGOMERY,K
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:08

SUBJECT'S LAST NAME: ESCAR

DATE/TIME OF TEST: 01/01/2016 04:34:52

Diagnostics Check		OK	04:28
Air Blank		0.000	04:29
Control Test		0.078	04:29
Air Blank		0.000	04:29
Subject Sample	# 1	0.106	04:30
Breath Volume		1.695	04:30
Air Blank		0.000	04:30
Air Blank		0.000	04:32
Subject Sample	# 2	0.108	04:33
Breath Volume		1.316	04:33
Air Blank		0.000	04:33
Control Test		0.079	04:34
Air Blank		0.000	04:34
Diagnostics Check		OK	04:34

Data Download Date/Time: 01/15/2016 09:08

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HORTON,REBECCA,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: KEY,M
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 19:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:16

SUBJECT'S LAST NAME: OLARU

DATE/TIME OF TEST: 01/01/2016 21:56:49

Diagnostics Check		OK	21:53	
Air Blank		0.000	21:53	
Control Test		0.079	21:53	
Air Blank		0.000	21:54	
Subject Sample	# 1		21:55	Subject Test Refused
Breath Volume		0.000	21:55	
Air Blank		0.000	21:55	
Control Test		0.080	21:56	
Air Blank		0.000	21:56	
Diagnostics Check		OK	21:56	

Data Download Date/Time: 01/15/2016 09:08

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HORTON,REBECCA,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FISETTE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: MAPLE

DATE/TIME OF TEST: 01/01/2016 22:32:14

Diagnostics Check		OK	22:25
Air Blank		0.000	22:26
Control Test		0.080	22:26
Air Blank		0.000	22:26
Subject Sample	# 1	0.195	22:27
Breath Volume		4.335	22:27
Air Blank		0.000	22:28
Air Blank		0.000	22:30
Subject Sample	# 2	0.195	22:30
Breath Volume		4.074	22:30
Air Blank		0.000	22:31
Control Test		0.079	22:31
Air Blank		0.000	22:32
Diagnostics Check		OK	22:32

Data Download Date/Time: 01/15/2016 09:08

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DASHER,JOSHUA,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: TERRELL,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:42

SUBJECT'S LAST NAME: NORRIS

DATE/TIME OF TEST: 01/01/2016 23:08:31

Diagnostics Check		OK	23:02
Air Blank		0.000	23:02
Control Test		0.079	23:03
Air Blank		0.000	23:03
Subject Sample	# 1	0.081	23:04
Breath Volume		2.039	23:04
Air Blank		0.000	23:04
Air Blank		0.000	23:06
Subject Sample	# 2	0.082	23:07
Breath Volume		2.121	23:07
Air Blank		0.000	23:07
Control Test		0.079	23:07
Air Blank		0.000	23:08
Diagnostics Check		OK	23:08

Data Download Date/Time: 01/15/2016 09:08

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: GARCIA,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:16

SUBJECT'S LAST NAME: MOLY

DATE/TIME OF TEST: 01/02/2016 00:50:15

Diagnostics Check		OK	00:47	
Air Blank		0.000	00:47	
Control Test		0.080	00:48	
Air Blank		0.000	00:48	
Subject Sample	# 1		00:48	Subject Test Refused
Breath Volume		0.000	00:48	
Air Blank		0.000	00:49	
Control Test		0.079	00:49	
Air Blank		0.000	00:50	
Diagnostics Check		OK	00:50	

Data Download Date/Time: 01/15/2016 09:08

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRIDGEMAN,JUSTIN,K
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FISETTE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 18:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:50

SUBJECT'S LAST NAME: FRANKLIN

DATE/TIME OF TEST: 01/02/2016 20:20:12

Diagnostics Check		OK	20:12
Air Blank		0.000	20:13
Control Test		0.080	20:13
Air Blank		0.000	20:14
Subject Sample	# 1	0.133	20:15
Breath Volume		1.390	20:15
Air Blank		0.000	20:15
Air Blank		0.000	20:17
Subject Sample	# 2	0.136	20:18
Breath Volume		1.355	20:18
Air Blank		0.000	20:19
Control Test		0.079	20:19
Air Blank		0.000	20:20
Diagnostics Check		OK	20:20

Data Download Date/Time: 01/15/2016 09:08

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRIDGEMAN,JUSTIN,K
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FISETTE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 19:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:46

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 01/04/2016 21:32:53

Diagnostics Check		OK	21:29	
Air Blank		0.000	21:30	
Control Test		0.080	21:30	
Air Blank		0.000	21:31	
Subject Sample	# 1		21:31	Subject Test Refused
Breath Volume		0.000	21:31	
Air Blank		0.000	21:31	
Control Test		0.080	21:32	
Air Blank		0.000	21:32	
Diagnostics Check		OK	21:32	

Data Download Date/Time: 01/15/2016 09:08

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DASHER,JOSHUA,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: GARCIA,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:26

SUBJECT'S LAST NAME: KING

DATE/TIME OF TEST: 01/04/2016 23:52:52

Diagnostics Check		OK	23:46
Air Blank		0.000	23:47
Control Test		0.081	23:47
Air Blank		0.000	23:47
Subject Sample	# 1	0.000	23:48
Breath Volume		1.160	23:48
Air Blank		0.000	23:48
Air Blank		0.000	23:50
Subject Sample	# 2	0.000	23:51
Breath Volume		1.351	23:51
Air Blank		0.000	23:51
Control Test		0.080	23:52
Air Blank		0.000	23:52
Diagnostics Check		OK	23:52

Data Download Date/Time: 01/15/2016 09:08

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DASHER,JOSHUA,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SCARBOROUGH,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:21

SUBJECT'S LAST NAME: RAY

DATE/TIME OF TEST: 01/05/2016 01:44:49

Diagnostics Check		OK	01:41	
Air Blank		0.000	01:42	
Control Test		0.080	01:42	
Air Blank		0.000	01:43	
Subject Sample	# 1		01:43	Subject Test Refused
Breath Volume		0.000	01:43	
Air Blank		0.000	01:43	
Control Test		0.081	01:44	
Air Blank		0.000	01:44	
Diagnostics Check		OK	01:44	

Data Download Date/Time: 01/15/2016 09:08

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DASHER,JOSHUA,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: TERRELL,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:59

SUBJECT'S LAST NAME: WICAL

DATE/TIME OF TEST: 01/05/2016 03:28:47

Diagnostics Check		OK	03:19
Air Blank		0.000	03:20
Control Test		0.080	03:20
Air Blank		0.000	03:21
Subject Sample	# 1	0.122	03:23
Breath Volume		3.457	03:23
Air Blank		0.000	03:23
Air Blank		0.000	03:25
Subject Sample	# 2	0.122	03:27
Breath Volume		2.011	03:27
Air Blank		0.000	03:27
Control Test		0.080	03:28
Air Blank		0.000	03:28
Diagnostics Check		OK	03:28

Data Download Date/Time: 01/15/2016 09:08

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEVERENZ,PATRICK,T
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BYRNE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:56

SUBJECT'S LAST NAME: ATTAWAY

DATE/TIME OF TEST: 01/06/2016 02:26:55

Diagnostics Check		OK	02:17
Air Blank		0.000	02:17
Control Test		0.080	02:18
Air Blank		0.000	02:18
Subject Sample	# 1	0.148	02:19
Breath Volume		1.128	02:19
Air Blank		0.000	02:19
Air Blank		0.000	02:21
Subject Sample	# 2	0.170	02:22
Breath Volume		2.933	02:22
Air Blank		0.000	02:23
Air Blank		0.000	02:25
Subject Sample	# 3	0.150	02:25
Breath Volume		1.562	02:25
Air Blank		0.000	02:26
Control Test		0.079	02:26
Air Blank		0.000	02:26
Diagnostics Check		OK	02:26

Data Download Date/Time: 01/15/2016 09:08

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEVERENZ,PATRICK,T
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BYRNE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:51

SUBJECT'S LAST NAME: GREEN JR

DATE/TIME OF TEST: 01/07/2016 01:17:36

Diagnostics Check		OK	01:11
Air Blank		0.000	01:11
Control Test		0.080	01:11
Air Blank		0.000	01:12
Subject Sample	# 1	0.173	01:12
Breath Volume		1.437	01:12
Air Blank		0.000	01:13
Air Blank		0.000	01:15
Subject Sample	# 2	0.167	01:16
Breath Volume		2.109	01:16
Air Blank		0.000	01:16
Control Test		0.080	01:17
Air Blank		0.000	01:17
Diagnostics Check		OK	01:17

Data Download Date/Time: 01/15/2016 09:08

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEVENS,DAVID,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: COLEMAN,F
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 18:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:04

SUBJECT'S LAST NAME: WILLIS

DATE/TIME OF TEST: 01/07/2016 20:26:59

Diagnostics Check		OK	20:24	
Air Blank		0.000	20:24	
Control Test		0.080	20:24	
Air Blank		0.000	20:25	
Subject Sample	# 1		20:25	Subject Test Refused
Breath Volume		0.000	20:25	
Air Blank		0.000	20:26	
Control Test		0.079	20:26	
Air Blank		0.000	20:26	
Diagnostics Check		OK	20:26	

Data Download Date/Time: 01/15/2016 09:08

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEVENS,DAVID,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: EANES,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:59

SUBJECT'S LAST NAME: REINOLDS

DATE/TIME OF TEST: 01/07/2016 22:30:05

Diagnostics Check		OK	22:19
Air Blank		0.000	22:20
Control Test		0.080	22:20
Air Blank		0.000	22:21
Subject Sample	# 1	0.059	22:24
Breath Volume		1.449	22:24
Air Blank		0.000	22:24
Air Blank		0.000	22:26
Subject Sample	# 2	0.061	22:28
Breath Volume		1.207	22:28
Air Blank		0.000	22:29
Control Test		0.079	22:29
Air Blank		0.000	22:30
Diagnostics Check		OK	22:30

Data Download Date/Time: 01/15/2016 09:08

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEVERENZ,PATRICK,T
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BYRNE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: MALOTT

DATE/TIME OF TEST: 01/08/2016 01:18:45

Diagnostics Check		OK	01:15	
Air Blank		0.000	01:16	
Control Test		0.081	01:16	
Air Blank		0.000	01:16	
Subject Sample	# 1		01:17	Subject Test Refused
Breath Volume		0.000	01:17	
Air Blank		0.000	01:17	
Control Test		0.080	01:18	
Air Blank		0.000	01:18	
Diagnostics Check		OK	01:18	

Data Download Date/Time: 01/15/2016 09:08

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBER,JOSEPH,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BROWN,T.A.
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: MEHMEDOVIC

DATE/TIME OF TEST: 01/09/2016 00:22:01

Diagnostics Check		OK	00:15
Air Blank		0.000	00:15
Control Test		0.080	00:15
Air Blank		0.000	00:16
Subject Sample	# 1	0.181	00:17
Breath Volume		2.199	00:17
Air Blank		0.000	00:18
Air Blank		0.000	00:19
Subject Sample	# 2	0.171	00:20
Breath Volume		1.378	00:20
Air Blank		0.000	00:21
Control Test		0.079	00:21
Air Blank		0.000	00:21
Diagnostics Check		OK	00:22

Data Download Date/Time: 01/15/2016 09:08

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MONTGOMERY,K
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:59

SUBJECT'S LAST NAME: ZUB

DATE/TIME OF TEST: 01/10/2016 00:30:20

Diagnostics Check		OK	00:22
Air Blank		0.000	00:23
Control Test		0.079	00:23
Air Blank		0.000	00:23
Subject Sample	# 1	0.120	00:25
Breath Volume		2.429	00:25
Air Blank		0.000	00:26
Air Blank		0.000	00:28
Subject Sample	# 2	0.116	00:28
Breath Volume		1.777	00:28
Air Blank		0.000	00:29
Control Test		0.078	00:29
Air Blank		0.000	00:30
Diagnostics Check		OK	00:30

Data Download Date/Time: 01/15/2016 09:08

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHITE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:18

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 01/10/2016 05:45:19

Diagnostics Check		OK	05:39
Air Blank		0.000	05:39
Control Test		0.080	05:39
Air Blank		0.000	05:40
Subject Sample	# 1	0.142	05:40
Breath Volume		3.992	05:40
Air Blank		0.000	05:41
Air Blank		0.000	05:43
Subject Sample	# 2	0.142	05:43
Breath Volume		4.851	05:43
Air Blank		0.000	05:44
Control Test		0.079	05:44
Air Blank		0.000	05:45
Diagnostics Check		OK	05:45

Data Download Date/Time: 01/15/2016 09:08

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRIDGEMAN,JUSTIN,K
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MEDINA,J
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 13:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:53

SUBJECT'S LAST NAME: MCGRIFF

DATE/TIME OF TEST: 01/10/2016 15:20:25

Diagnostics Check		OK	15:13
Air Blank		0.000	15:14
Control Test		0.079	15:14
Air Blank		0.000	15:15
Subject Sample	# 1	0.025	15:15
Breath Volume		1.742	15:15
Air Blank		0.000	15:16
Air Blank		0.000	15:18
Subject Sample	# 2	0.024	15:18
Breath Volume		1.390	15:18
Air Blank		0.000	15:19
Control Test		0.079	15:19
Air Blank		0.000	15:20
Diagnostics Check		OK	15:20

Data Download Date/Time: 01/15/2016 09:08

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRIDGEMAN,JUSTIN,K
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: GARCIA,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 19:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:02

SUBJECT'S LAST NAME: MESSERSCHMIDT

DATE/TIME OF TEST: 01/10/2016 21:29:51

Diagnostics Check		OK	21:22
Air Blank		0.000	21:22
Control Test		0.080	21:23
Air Blank		0.000	21:23
Subject Sample	# 1	0.218	21:24
Breath Volume		2.226	21:24
Air Blank		0.000	21:25
Air Blank		0.000	21:27
Subject Sample	# 2	0.215	21:28
Breath Volume		2.511	21:28
Air Blank		0.000	21:28
Control Test		0.080	21:29
Air Blank		0.000	21:29
Diagnostics Check		OK	21:29

Data Download Date/Time: 01/15/2016 09:08

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DASHER,JOSHUA,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MONTGOMERY,K
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 01/10/2016 23:47:06

Diagnostics Check		OK	23:40
Air Blank		0.000	23:41
Control Test		0.080	23:41
Air Blank		0.000	23:42
Subject Sample	# 1	0.174	23:42
Breath Volume		2.496	23:42
Air Blank		0.000	23:43
Air Blank		0.000	23:45
Subject Sample	# 2	0.170	23:45
Breath Volume		1.574	23:45
Air Blank		0.000	23:46
Control Test		0.080	23:46
Air Blank		0.000	23:47
Diagnostics Check		OK	23:47

Data Download Date/Time: 01/15/2016 09:08

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DASHER,JOSHUA,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WALL,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: BAILEY

DATE/TIME OF TEST: 01/11/2016 00:50:12

Diagnostics Check		OK	00:47	
Air Blank		0.000	00:47	
Control Test		0.080	00:48	
Air Blank		0.000	00:48	
Subject Sample	# 1		00:48	Subject Test Refused
Breath Volume		0.000	00:48	
Air Blank		0.000	00:49	
Control Test		0.080	00:49	
Air Blank		0.000	00:50	
Diagnostics Check		OK	00:50	

Data Download Date/Time: 01/15/2016 09:08

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DASHER,JOSHUA,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MONTGOMERY,K
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: SOLIS-MADRIGAL

DATE/TIME OF TEST: 01/11/2016 02:59:54

Diagnostics Check		OK	02:56	
Air Blank		0.000	02:57	
Control Test		0.080	02:57	
Air Blank		0.000	02:58	
Subject Sample	# 1		02:58	Subject Test Refused
Breath Volume		0.000	02:58	
Air Blank		0.000	02:58	
Control Test		0.081	02:59	
Air Blank		0.000	02:59	
Diagnostics Check		OK	02:59	

Data Download Date/Time: 01/15/2016 09:08

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBER,JOSEPH,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: TERRELL,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:31

SUBJECT'S LAST NAME: RILEY

DATE/TIME OF TEST: 01/11/2016 23:54:35

Diagnostics Check		OK	23:51	
Air Blank		0.000	23:51	
Control Test		0.081	23:52	
Air Blank		0.000	23:52	
Subject Sample	# 1		23:53	Subject Test Refused
Breath Volume		0.000	23:53	
Air Blank		0.000	23:53	
Control Test		0.080	23:53	
Air Blank		0.000	23:54	
Diagnostics Check		OK	23:54	

Data Download Date/Time: 01/15/2016 09:08

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBER,JOSEPH,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: TERRELL,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:03

SUBJECT'S LAST NAME: MIDGLEY

DATE/TIME OF TEST: 01/12/2016 03:29:58

Diagnostics Check		OK	03:23
Air Blank		0.000	03:24
Control Test		0.081	03:24
Air Blank		0.000	03:24
Subject Sample	# 1	0.177	03:25
Breath Volume		3.355	03:25
Air Blank		0.000	03:26
Air Blank		0.000	03:27
Subject Sample	# 2	0.171	03:28
Breath Volume		4.472	03:28
Air Blank		0.000	03:29
Control Test		0.080	03:29
Air Blank		0.000	03:29
Diagnostics Check		OK	03:29

Data Download Date/Time: 01/15/2016 09:08

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRIDGEMAN,JUSTIN,K
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: TERRELL,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 19:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:39

SUBJECT'S LAST NAME: PELLETIER

DATE/TIME OF TEST: 01/13/2016 21:06:25

Diagnostics Check		OK	21:00
Air Blank		0.000	21:00
Control Test		0.081	21:00
Air Blank		0.000	21:01
Subject Sample	# 1	0.337	21:01
Breath Volume		2.738	21:01
Air Blank		0.000	21:02
Air Blank		0.000	21:04
Subject Sample	# 2	0.337	21:04
Breath Volume		2.683	21:04
Air Blank		0.000	21:05
Control Test		0.080	21:05
Air Blank		0.000	21:06
Diagnostics Check		OK	21:06

Data Download Date/Time: 01/15/2016 09:08

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HORTON,REBECCA,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MARRERO,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/30/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 17:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:35

SUBJECT'S LAST NAME: KENYON

DATE/TIME OF TEST: 01/14/2016 19:02:29

Diagnostics Check		OK	18:59	
Air Blank		0.000	18:59	
Control Test		0.080	18:59	
Air Blank		0.000	19:00	
Subject Sample	# 1		19:01	Subject Test Refused
Breath Volume		0.000	19:01	
Air Blank		0.000	19:01	
Control Test		0.080	19:01	
Air Blank		0.000	19:02	
Diagnostics Check		OK	19:02	

Data Download Date/Time: 01/15/2016 09:08

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRIDGEMAN,JUSTIN,K
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BYRNE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:41

SUBJECT'S LAST NAME: MCDANIEL

DATE/TIME OF TEST: 01/15/2016 23:12:47

Diagnostics Check		OK	23:02
Air Blank		0.000	23:02
Control Test		0.079	23:03
Air Blank		0.000	23:03
Subject Sample	# 1	0.202	23:04
Breath Volume		1.324	23:04
Air Blank		0.000	23:05
Air Blank		0.000	23:07
Subject Sample	# 2	0.223	23:08
Breath Volume		2.601	23:08
Air Blank		0.000	23:08
Air Blank		0.000	23:10
Subject Sample	# 3	0.221	23:11
Breath Volume		2.820	23:11
Air Blank		0.000	23:11
Control Test		0.078	23:12
Air Blank		0.000	23:12
Diagnostics Check		OK	23:12

Data Download Date/Time: 01/29/2016 08:48

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEVERENZ,PATRICK,T
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: EANES,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:16

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 01/16/2016 01:43:13

Diagnostics Check		OK	01:36
Air Blank		0.000	01:37
Control Test		0.079	01:37
Air Blank		0.000	01:38
Subject Sample	# 1	0.100	01:38
Breath Volume		3.882	01:38
Air Blank		0.000	01:39
Air Blank		0.000	01:41
Subject Sample	# 2	0.085	01:41
Breath Volume		1.242	01:41
Air Blank		0.000	01:42
Control Test		0.079	01:42
Air Blank		0.000	01:43
Diagnostics Check		OK	01:43

Data Download Date/Time: 01/29/2016 08:48

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEVERENZ,PATRICK,T
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BYRNE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:42

SUBJECT'S LAST NAME: STEWART

DATE/TIME OF TEST: 01/16/2016 04:05:07

Diagnostics Check		OK	04:02	
Air Blank		0.000	04:02	
Control Test		0.080	04:02	
Air Blank		0.000	04:03	
Subject Sample	# 1		04:03	Subject Test Refused
Breath Volume		0.000	04:03	
Air Blank		0.000	04:04	
Control Test		0.080	04:04	
Air Blank		0.000	04:05	
Diagnostics Check		OK	04:05	

Data Download Date/Time: 01/29/2016 08:48

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEVERENZ,PATRICK,T
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CENSON,T
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 01/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 05:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:24

SUBJECT'S LAST NAME: REYNOLDS

DATE/TIME OF TEST: 01/16/2016 06:50:59

Diagnostics Check		OK	06:44
Air Blank		0.000	06:45
Control Test		0.080	06:45
Air Blank		0.000	06:46
Subject Sample	# 1	0.000	06:46
Breath Volume		4.753	06:46
Air Blank		0.000	06:47
Air Blank		0.000	06:49
Subject Sample	# 2	0.000	06:49
Breath Volume		4.246	06:49
Air Blank		0.000	06:50
Control Test		0.079	06:50
Air Blank		0.000	06:50
Diagnostics Check		OK	06:50

Data Download Date/Time: 01/29/2016 08:48

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEVERENZ,PATRICK,T
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BAI,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:09

SUBJECT'S LAST NAME: FLORES MORENO

DATE/TIME OF TEST: 01/17/2016 03:36:22

Diagnostics Check		OK	03:29
Air Blank		0.000	03:30
Control Test		0.079	03:30
Air Blank		0.000	03:31
Subject Sample	# 1	0.187	03:31
Breath Volume		1.605	03:31
Air Blank		0.000	03:32
Air Blank		0.000	03:34
Subject Sample	# 2	0.187	03:34
Breath Volume		2.183	03:34
Air Blank		0.000	03:35
Control Test		0.078	03:35
Air Blank		0.000	03:36
Diagnostics Check		OK	03:36

Data Download Date/Time: 01/29/2016 08:48

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLEY,SHANNON,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BARRIER,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 05:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:04

SUBJECT'S LAST NAME: SCHIVELBEIN

DATE/TIME OF TEST: 01/17/2016 07:52:17

Diagnostics Check		OK	07:49	
Air Blank		0.000	07:49	
Control Test		0.079	07:50	
Air Blank		0.000	07:50	
Subject Sample	# 1		07:50	Subject Test Refused
Breath Volume		0.000	07:50	
Air Blank		0.000	07:51	
Control Test		0.079	07:51	
Air Blank		0.000	07:52	
Diagnostics Check		OK	07:52	

Data Download Date/Time: 01/29/2016 08:48

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEVERENZ,PATRICK,T
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BROWN,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 01/18/2016 00:34:51

Diagnostics Check		OK	00:32	
Air Blank		0.000	00:32	
Control Test		0.080	00:32	
Air Blank		0.000	00:33	
Subject Sample	# 1		00:33	Subject Test Refused
Breath Volume		0.000	00:33	
Air Blank		0.000	00:33	
Control Test		0.080	00:34	
Air Blank		0.000	00:34	
Diagnostics Check		OK	00:34	

Data Download Date/Time: 01/29/2016 08:48

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEVERENZ,PATRICK,T
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BRYANT,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:46

SUBJECT'S LAST NAME: PAULUS

DATE/TIME OF TEST: 01/18/2016 23:13:11

Diagnostics Check		OK	23:07
Air Blank		0.000	23:07
Control Test		0.080	23:07
Air Blank		0.000	23:08
Subject Sample	# 1	0.279	23:08
Breath Volume		5.062	23:08
Air Blank		0.000	23:09
Air Blank		0.000	23:11
Subject Sample	# 2	0.283	23:11
Breath Volume		4.601	23:11
Air Blank		0.000	23:12
Control Test		0.080	23:12
Air Blank		0.000	23:13
Diagnostics Check		OK	23:13

Data Download Date/Time: 01/29/2016 08:48

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEVERENZ,PATRICK,T
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BARTILUCCI,D
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:26

SUBJECT'S LAST NAME: QUAINTANCE

DATE/TIME OF TEST: 01/21/2016 23:50:08

Diagnostics Check		OK	23:47	
Air Blank		0.000	23:47	
Control Test		0.079	23:47	
Air Blank		0.000	23:48	
Subject Sample	# 1		23:48	Subject Test Refused
Breath Volume		0.000	23:48	
Air Blank		0.000	23:49	
Control Test		0.080	23:49	
Air Blank		0.000	23:50	
Diagnostics Check		OK	23:50	

Data Download Date/Time: 01/29/2016 08:48

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEVERENZ,PATRICK,T
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: TERRELL,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: LEWIS

DATE/TIME OF TEST: 01/22/2016 02:30:53

Diagnostics Check		OK	02:28	
Air Blank		0.000	02:28	
Control Test		0.080	02:28	
Air Blank		0.000	02:29	
Subject Sample	# 1		02:29	Subject Test Refused
Breath Volume		0.000	02:29	
Air Blank		0.000	02:29	
Control Test		0.080	02:30	
Air Blank		0.000	02:30	
Diagnostics Check		OK	02:30	

Data Download Date/Time: 01/29/2016 08:48

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRIDGEMAN,JUSTIN,K
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BRYANT,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 19:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:55

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 01/25/2016 21:19:59

Diagnostics Check		OK	21:16	
Air Blank		0.000	21:16	
Control Test		0.080	21:16	
Air Blank		0.000	21:17	
Subject Sample	# 1		21:18	Subject Test Refused
Breath Volume		0.835	21:18	
Air Blank		0.000	21:19	
Control Test		0.080	21:19	
Air Blank		0.000	21:19	
Diagnostics Check		OK	21:19	

Data Download Date/Time: 01/29/2016 08:48

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEVERENZ,PATRICK,T
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: COLEMAN,F
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:57

SUBJECT'S LAST NAME: HANKS

DATE/TIME OF TEST: 01/27/2016 03:25:37

Diagnostics Check		OK	03:18
Air Blank		0.000	03:19
Control Test		0.080	03:19
Air Blank		0.000	03:19
Subject Sample	# 1	0.152	03:21
Breath Volume		1.246	03:21
Air Blank		0.000	03:21
Air Blank		0.000	03:23
Subject Sample	# 2	0.160	03:24
Breath Volume		1.546	03:24
Air Blank		0.000	03:24
Control Test		0.079	03:25
Air Blank		0.000	03:25
Diagnostics Check		OK	03:25

Data Download Date/Time: 01/29/2016 08:48

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FISETTE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:14

SUBJECT'S LAST NAME: PAALA

DATE/TIME OF TEST: 01/29/2016 01:41:49

Diagnostics Check		OK	01:35
Air Blank		0.000	01:35
Control Test		0.080	01:36
Air Blank		0.000	01:36
Subject Sample	# 1	0.138	01:37
Breath Volume		3.234	01:37
Air Blank		0.000	01:37
Air Blank		0.000	01:39
Subject Sample	# 2	0.139	01:40
Breath Volume		3.496	01:40
Air Blank		0.000	01:40
Control Test		0.080	01:41
Air Blank		0.000	01:41
Diagnostics Check		OK	01:41

Data Download Date/Time: 01/29/2016 08:48

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001109

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,WILLIAM,T
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SMITH,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/03/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 16014080A2
DRY GAS EXPIRY DATE: 07/05/2016
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:51

SUBJECT'S LAST NAME: TABORA

DATE/TIME OF TEST: 01/02/2016 03:18:08

Diagnostics Check		OK	03:11
Air Blank		0.000	03:11
Control Test		0.078	03:12
Air Blank		0.000	03:12
Subject Sample	# 1	0.123	03:13
Breath Volume		2.109	03:13
Air Blank		0.000	03:14
Air Blank		0.000	03:16
Subject Sample	# 2	0.116	03:16
Breath Volume		1.609	03:16
Air Blank		0.000	03:17
Control Test		0.078	03:17
Air Blank		0.000	03:18
Diagnostics Check		OK	03:18

Data Download Date/Time: 01/15/2016 22:36

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001109

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOLAN,NICHOLAS,C
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: DOLAN,NICHOLAS
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/03/2015

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 16014080A2
DRY GAS EXPIRY DATE: 07/05/2016
VIOLATION CODE: DUI
ARREST TIME: 00:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: ELI

DATE/TIME OF TEST: 01/03/2016 01:02:45

Diagnostics Check		OK	00:55
Air Blank		0.000	00:55
Control Test		0.079	00:56
Air Blank		0.000	00:56
Subject Sample	# 1	0.171	00:58
Breath Volume		3.464	00:58
Air Blank		0.000	00:58
Air Blank		0.000	01:00
Subject Sample	# 2	0.155	01:01
Breath Volume		1.835	01:01
Air Blank		0.000	01:01
Control Test		0.076	01:02
Air Blank		0.000	01:02
Diagnostics Check		OK	01:02

Data Download Date/Time: 01/15/2016 22:36

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001109

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANGELICCHI,JOSEPH,S
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: ANGELICCHI,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/03/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 16014080A2
DRY GAS EXPIRY DATE: 07/05/2016
VIOLATION CODE: DUI
ARREST TIME: 00:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: CREIGHTON

DATE/TIME OF TEST: 01/03/2016 01:20:53

Diagnostics Check		OK	01:18	
Air Blank		0.000	01:18	
Control Test		0.077	01:18	
Air Blank		0.000	01:19	
Subject Sample	# 1		01:19	Subject Test Refused
Breath Volume		0.000	01:19	
Air Blank		0.000	01:19	
Control Test		0.077	01:20	
Air Blank		0.000	01:20	
Diagnostics Check		OK	01:20	

Data Download Date/Time: 01/15/2016 22:36

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001109

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,ALAIN,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: HERNANDEZ,ALAIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/03/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 16014080A2
DRY GAS EXPIRY DATE: 07/05/2016
VIOLATION CODE: DUI
ARREST TIME: 02:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:12

SUBJECT'S LAST NAME: MOREIRA-VENEGAS

DATE/TIME OF TEST: 01/03/2016 02:44:02

Diagnostics Check		OK	02:32	
Air Blank		0.000	02:33	
Control Test		0.078	02:33	
Air Blank		0.000	02:34	
Subject Sample	# 1	0.158	02:36	
Breath Volume		1.148	02:36	
Air Blank		0.000	02:37	
Air Blank		0.000	02:39	
Subject Sample	# 2		02:42	Subject Test Refused
Breath Volume		0.175	02:42	
Air Blank		0.000	02:43	
Control Test		0.077	02:43	
Air Blank		0.000	02:43	
Diagnostics Check		OK	02:44	

Data Download Date/Time: 01/15/2016 22:36

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001109

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLIS JR,MICHAEL,A
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: HOLLIS JR,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/03/2015

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 16014080A2
DRY GAS EXPIRY DATE: 07/05/2016
VIOLATION CODE: DUI
ARREST TIME: 02:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:54

SUBJECT'S LAST NAME: TAGO

DATE/TIME OF TEST: 01/07/2016 02:25:54

Diagnostics Check		OK	02:17
Air Blank		0.000	02:18
Control Test		0.078	02:18
Air Blank		0.000	02:19
Subject Sample	# 1	0.183	02:21
Breath Volume		1.511	02:21
Air Blank		0.000	02:21
Air Blank		0.000	02:23
Subject Sample	# 2	0.187	02:24
Breath Volume		2.046	02:24
Air Blank		0.000	02:25
Control Test		0.077	02:25
Air Blank		0.000	02:25
Diagnostics Check		OK	02:25

Data Download Date/Time: 01/15/2016 22:37

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001109

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLIS JR,MICHAEL,A
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: HOLLIS JR,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/03/2015

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 16014080A2
DRY GAS EXPIRY DATE: 07/05/2016
VIOLATION CODE: DUI
ARREST TIME: 20:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:43

SUBJECT'S LAST NAME: PENVOSE

DATE/TIME OF TEST: 01/07/2016 21:24:21

Diagnostics Check		OK	21:08	
Air Blank		0.000	21:08	
Control Test		0.079	21:08	
Air Blank		0.000	21:09	
Subject Sample	# 1	0.100	21:12	Volume Not Met
Breath Volume		1.281	21:12	
Air Blank		0.000	21:13	
Air Blank		0.000	21:14	
Subject Sample	# 2	0.102	21:17	
Breath Volume		1.125	21:17	
Air Blank		0.000	21:17	
Air Blank		0.000	21:19	
Subject Sample	# 3	0.103	21:22	Volume Not Met
Breath Volume		0.109	21:22	
Air Blank		0.000	21:23	
Control Test		0.076	21:23	
Air Blank		0.000	21:24	
Diagnostics Check		OK	21:24	

Data Download Date/Time: 01/15/2016 22:37

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001109

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLIS JR,MICHAEL,A
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: HOLLIS JR,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/03/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 16014080A2
DRY GAS EXPIRY DATE: 07/05/2016
VIOLATION CODE: DUI
ARREST TIME: 20:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:43

SUBJECT'S LAST NAME: PENVOSE

DATE/TIME OF TEST: 01/07/2016 21:37:16

Diagnostics Check		OK	21:26	
Air Blank		0.000	21:27	
Control Test		0.077	21:27	
Air Blank		0.000	21:28	
Subject Sample	# 1	0.090	21:31	Volume Not Met
Breath Volume		0.148	21:31	
Air Blank		0.000	21:32	
Air Blank		0.000	21:33	
Subject Sample	# 2		21:35	Subject Test Refused
Breath Volume		0.332	21:35	
Air Blank		0.000	21:36	
Control Test		0.077	21:36	
Air Blank		0.000	21:37	
Diagnostics Check		OK	21:37	

Data Download Date/Time: 01/15/2016 22:37

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001109

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANGELICCHI,JOSEPH,S
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: ANGELICCHI,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/03/2015

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 16014080A2
DRY GAS EXPIRY DATE: 07/05/2016
VIOLATION CODE: DUI
ARREST TIME: 23:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: SMALL

DATE/TIME OF TEST: 01/09/2016 23:44:57

Diagnostics Check		OK	23:38
Air Blank		0.000	23:39
Control Test		0.078	23:39
Air Blank		0.000	23:39
Subject Sample	# 1	0.299	23:40
Breath Volume		1.230	23:40
Air Blank		0.000	23:41
Air Blank		0.000	23:43
Subject Sample	# 2	0.289	23:43
Breath Volume		1.136	23:43
Air Blank		0.000	23:44
Control Test		0.078	23:44
Air Blank		0.000	23:44
Diagnostics Check		OK	23:44

Data Download Date/Time: 01/15/2016 22:37

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001109

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLIS JR,MICHAEL,A
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: HOLLIS JR,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/03/2015

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 16014080A2
DRY GAS EXPIRY DATE: 07/05/2016
VIOLATION CODE: DUI
ARREST TIME: 21:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:59

SUBJECT'S LAST NAME: MERKWA

DATE/TIME OF TEST: 01/14/2016 21:34:41

Diagnostics Check		OK	21:28
Air Blank		0.000	21:28
Control Test		0.079	21:28
Air Blank		0.000	21:29
Subject Sample	# 1	0.072	21:29
Breath Volume		2.156	21:29
Air Blank		0.000	21:30
Air Blank		0.000	21:32
Subject Sample	# 2	0.074	21:33
Breath Volume		2.285	21:33
Air Blank		0.000	21:33
Control Test		0.078	21:34
Air Blank		0.000	21:34
Diagnostics Check		OK	21:34

Data Download Date/Time: 01/15/2016 22:37

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001109

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANGELICCHI,JOSEPH,S
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: ANGELICCHI,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/15/2016

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 16014080A2
DRY GAS EXPIRY DATE: 07/05/2016
VIOLATION CODE: DUI
ARREST TIME: 23:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:11

SUBJECT'S LAST NAME: CHERIAN

DATE/TIME OF TEST: 01/16/2016 23:38:54

Diagnostics Check		OK	23:31
Air Blank		0.000	23:32
Control Test		0.079	23:32
Air Blank		0.000	23:33
Subject Sample	# 1	0.107	23:34
Breath Volume		1.937	23:34
Air Blank		0.000	23:34
Air Blank		0.000	23:36
Subject Sample	# 2	0.107	23:37
Breath Volume		1.882	23:37
Air Blank		0.000	23:38
Control Test		0.079	23:38
Air Blank		0.000	23:38
Diagnostics Check		OK	23:38

Data Download Date/Time: 02/24/2016 19:47

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001110

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADDISON,RAYMOND,E
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: SANTANA,YUDIEL
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 12/03/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1878845
DRY GAS EXPIRY DATE: 06/22/2018
VIOLATION CODE: DUI
ARREST TIME: 02:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:32

SUBJECT'S LAST NAME: RUEDA

DATE/TIME OF TEST: 01/11/2016 02:58:19

Diagnostics Check		OK	02:55	
Air Blank		0.000	02:55	
Control Test		0.080	02:56	
Air Blank		0.000	02:56	
Subject Sample	# 1		02:56	Subject Test Refused
Breath Volume		0.000	02:56	
Air Blank		0.000	02:57	
Control Test		0.080	02:57	
Air Blank		0.000	02:58	
Diagnostics Check		OK	02:58	

Data Download Date/Time: 01/22/2016 10:51

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001110

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADDISON,RAYMOND,E
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: OCEGUERA,YASMANI
ARRESTING OFFICER AGENCY: Florida Intern Unv
LAST AGENCY INSPECTION: 12/03/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1878845
DRY GAS EXPIRY DATE: 06/22/2018
VIOLATION CODE: DUI
ARREST TIME: 06:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:11

SUBJECT'S LAST NAME: BARBOSA TRUJILLO

DATE/TIME OF TEST: 01/17/2016 06:42:32

Diagnostics Check		OK	06:36
Air Blank		0.000	06:36
Control Test		0.079	06:36
Air Blank		0.000	06:37
Subject Sample	# 1	0.156	06:37
Breath Volume		2.414	06:37
Air Blank		0.000	06:38
Air Blank		0.000	06:40
Subject Sample	# 2	0.156	06:40
Breath Volume		3.156	06:40
Air Blank		0.000	06:41
Control Test		0.079	06:41
Air Blank		0.000	06:42
Diagnostics Check		OK	06:42

Data Download Date/Time: 01/22/2016 10:51

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001110

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADDISON,RAYMOND,E
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: CASTILLO,RICARDO
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 01/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1878845
DRY GAS EXPIRY DATE: 06/22/2018
VIOLATION CODE: DUI
ARREST TIME: 03:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:19

SUBJECT'S LAST NAME: VEGA

DATE/TIME OF TEST: 01/25/2016 03:57:09

Diagnostics Check		OK	03:53	
Air Blank		0.000	03:54	
Control Test		0.082	03:54	
Air Blank		0.000	03:55	
Subject Sample	# 1		03:55	Subject Test Refused
Breath Volume		0.000	03:55	
Air Blank		0.000	03:56	
Control Test		0.082	03:56	
Air Blank		0.000	03:57	
Diagnostics Check		OK	03:57	

Data Download Date/Time: 02/27/2016 11:24

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001111

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICHARDSON,LOGAN,M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: RICHARDSON,LOGAN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 12/08/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 17:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:14

SUBJECT'S LAST NAME: DONNELL

DATE/TIME OF TEST: 01/04/2016 18:42:23

Diagnostics Check		OK	18:35
Air Blank		0.000	18:36
Control Test		0.080	18:36
Air Blank		0.000	18:37
Subject Sample	# 1	0.120	18:37
Breath Volume		1.953	18:37
Air Blank		0.000	18:38
Air Blank		0.000	18:40
Subject Sample	# 2	0.121	18:40
Breath Volume		2.441	18:40
Air Blank		0.000	18:41
Control Test		0.082	18:41
Air Blank		0.000	18:42
Diagnostics Check		OK	18:42

Data Download Date/Time: 01/27/2016 19:24

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001111

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICHARDSON,LOGAN,M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: RICHARDSON,LOGAN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 12/08/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 12:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:43

SUBJECT'S LAST NAME: ROYAL

DATE/TIME OF TEST: 01/05/2016 14:09:30

Diagnostics Check		OK	14:03
Air Blank		0.000	14:03
Control Test		0.080	14:04
Air Blank		0.000	14:04
Subject Sample	# 1	0.282	14:05
Breath Volume		2.921	14:05
Air Blank		0.000	14:05
Air Blank		0.000	14:07
Subject Sample	# 2	0.262	14:08
Breath Volume		2.339	14:08
Air Blank		0.000	14:08
Control Test		0.077	14:08
Air Blank		0.000	14:09
Diagnostics Check		OK	14:09

Data Download Date/Time: 01/27/2016 19:24

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001111

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAILEY,RAYMOND,F
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BAILEY,RAYMOND
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 12/08/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 15:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:05

SUBJECT'S LAST NAME: BIELA

DATE/TIME OF TEST: 01/20/2016 16:36:18

Diagnostics Check		OK	16:25
Air Blank		0.000	16:26
Control Test		0.082	16:26
Air Blank		0.000	16:26
Subject Sample	# 1	0.168	16:27
Breath Volume		2.535	16:27
Air Blank		0.000	16:27
Air Blank		0.000	16:29
Subject Sample	# 2	0.146	16:31
Breath Volume		1.734	16:31
Air Blank		0.000	16:32
Air Blank		0.000	16:34
Subject Sample	# 3	0.158	16:34
Breath Volume		2.000	16:34
Air Blank		0.000	16:35
Control Test		0.080	16:35
Air Blank		0.000	16:36
Diagnostics Check		OK	16:36

Data Download Date/Time: 01/27/2016 19:24

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001111

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SOLOMON,J,R
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: SOLOMON,J
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 12/08/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:45

SUBJECT'S LAST NAME: CAUDILL

DATE/TIME OF TEST: 01/23/2016 05:24:20

Diagnostics Check		OK	05:20	
Air Blank		0.000	05:21	
Control Test		0.081	05:21	
Air Blank		0.000	05:22	
Subject Sample	# 1		05:22	Subject Test Refused
Breath Volume		0.000	05:22	
Air Blank		0.000	05:23	
Control Test		0.082	05:23	
Air Blank		0.000	05:24	
Diagnostics Check		OK	05:24	

Data Download Date/Time: 01/27/2016 19:24

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001111

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GILL,BRIAN,K
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: GILL,BRIAN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 12/08/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 16:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:00

SUBJECT'S LAST NAME: LIEBERMAN

DATE/TIME OF TEST: 01/26/2016 17:27:41

Diagnostics Check		OK	17:20
Air Blank		0.000	17:20
Control Test		0.081	17:21
Air Blank		0.000	17:21
Subject Sample	# 1	0.176	17:22
Breath Volume		1.628	17:22
Air Blank		0.000	17:23
Air Blank		0.000	17:25
Subject Sample	# 2	0.171	17:26
Breath Volume		1.554	17:26
Air Blank		0.000	17:26
Control Test		0.081	17:27
Air Blank		0.000	17:27
Diagnostics Check		OK	17:27

Data Download Date/Time: 01/27/2016 19:24

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001112

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WENZEL,THOMAS,J
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: BEHRENSHOUSER,RICHARD
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 12/28/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1686121
DRY GAS EXPIRY DATE: 05/21/2017
VIOLATION CODE: DUI
ARREST TIME: 00:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: LEE

DATE/TIME OF TEST: 01/02/2016 02:17:24

Diagnostics Check		OK	02:11
Air Blank		0.000	02:11
Control Test		0.082	02:12
Air Blank		0.000	02:12
Subject Sample	# 1	0.161	02:13
Breath Volume		3.394	02:13
Air Blank		0.000	02:13
Air Blank		0.000	02:15
Subject Sample	# 2	0.162	02:15
Breath Volume		3.347	02:15
Air Blank		0.000	02:16
Control Test		0.082	02:16
Air Blank		0.000	02:17
Diagnostics Check		OK	02:17

Data Download Date/Time: 01/26/2016 04:00

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001112

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BEHRENSHOUSER,RICHARD,I
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: BEHRENSHOUSER,RICHARD
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 12/28/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1686121
DRY GAS EXPIRY DATE: 05/21/2017
VIOLATION CODE: DUI
ARREST TIME: 01:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: HORAN

DATE/TIME OF TEST: 01/09/2016 03:19:39

Diagnostics Check		OK	03:10
Air Blank		0.000	03:10
Control Test		0.083	03:11
Air Blank		0.000	03:11
Subject Sample	# 1	0.228	03:11
Breath Volume		2.246	03:11
Air Blank		0.000	03:12
Air Blank		0.000	03:14
Subject Sample	# 2	0.206	03:15
Breath Volume		1.867	03:15
Air Blank		0.000	03:15
Air Blank		0.000	03:17
Subject Sample	# 3	0.207	03:18
Breath Volume		1.812	03:18
Air Blank		0.000	03:18
Control Test		0.081	03:19
Air Blank		0.000	03:19
Diagnostics Check		OK	03:19

Data Download Date/Time: 01/26/2016 04:01

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001112

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HITCHCOCK,WILLIAM,D
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: HITCHCOCK,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 12/28/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1686121
DRY GAS EXPIRY DATE: 05/21/2017
VIOLATION CODE: DUI
ARREST TIME: 03:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:16

SUBJECT'S LAST NAME: BULL

DATE/TIME OF TEST: 01/11/2016 03:50:38

Diagnostics Check		OK	03:44
Air Blank		0.000	03:45
Control Test		0.082	03:45
Air Blank		0.000	03:45
Subject Sample	# 1	0.203	03:46
Breath Volume		2.085	03:46
Air Blank		0.000	03:46
Air Blank		0.000	03:48
Subject Sample	# 2	0.211	03:49
Breath Volume		2.085	03:49
Air Blank		0.000	03:49
Control Test		0.083	03:50
Air Blank		0.000	03:50
Diagnostics Check		OK	03:50

Data Download Date/Time: 01/26/2016 04:01

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001112

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BEHRENSHOUSER,RICHARD,I
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: BEHRENSHOUSER,RICHARD
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 12/28/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1686121
DRY GAS EXPIRY DATE: 05/21/2017
VIOLATION CODE: DUI
ARREST TIME: 21:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:53

SUBJECT'S LAST NAME: HEPLER

DATE/TIME OF TEST: 01/11/2016 22:20:17

Diagnostics Check		OK	22:13
Air Blank		0.000	22:14
Control Test		0.084	22:14
Air Blank		0.000	22:15
Subject Sample	# 1	0.280	22:15
Breath Volume		2.722	22:15
Air Blank		0.000	22:16
Air Blank		0.000	22:18
Subject Sample	# 2	0.275	22:18
Breath Volume		2.949	22:18
Air Blank		0.000	22:19
Control Test		0.083	22:19
Air Blank		0.000	22:20
Diagnostics Check		OK	22:20

Data Download Date/Time: 01/26/2016 04:01

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001112

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOLFF,ROSE,M
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: WOLFF,ROSE
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 12/28/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1686121
DRY GAS EXPIRY DATE: 05/21/2017
VIOLATION CODE: DUI
ARREST TIME: 05:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:00

SUBJECT'S LAST NAME: COMO

DATE/TIME OF TEST: 01/13/2016 06:42:06

Diagnostics Check		OK	06:34
Air Blank		0.000	06:34
Control Test		0.083	06:34
Air Blank		0.000	06:35
Subject Sample	# 1	0.093	06:36
Breath Volume		1.359	06:36
Air Blank		0.000	06:36
Air Blank		0.000	06:38
Subject Sample	# 2	0.089	06:40
Breath Volume		1.214	06:40
Air Blank		0.000	06:41
Control Test		0.084	06:41
Air Blank		0.000	06:42
Diagnostics Check		OK	06:42

Data Download Date/Time: 01/26/2016 04:01

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001112

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BEHRENSHOUSER,RICHARD,I
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: BEHRENSHOUSER,RICHARD
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 12/28/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1686121
DRY GAS EXPIRY DATE: 05/21/2017
VIOLATION CODE: DUI
ARREST TIME: 01:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: MADERA IRIZARRY

DATE/TIME OF TEST: 01/16/2016 02:22:51

Diagnostics Check		OK	02:16
Air Blank		0.000	02:16
Control Test		0.083	02:17
Air Blank		0.000	02:17
Subject Sample	# 1	0.095	02:18
Breath Volume		1.738	02:18
Air Blank		0.000	02:18
Air Blank		0.000	02:20
Subject Sample	# 2	0.092	02:21
Breath Volume		1.527	02:21
Air Blank		0.000	02:21
Control Test		0.083	02:22
Air Blank		0.000	02:22
Diagnostics Check		OK	02:22

Data Download Date/Time: 01/26/2016 04:01

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001112

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HITCHCOCK,WILLIAM,D
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: HITCHCOCK,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 12/28/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1686121
DRY GAS EXPIRY DATE: 05/21/2017
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: MOODY

DATE/TIME OF TEST: 01/17/2016 03:43:24

Diagnostics Check		OK	03:37
Air Blank		0.000	03:37
Control Test		0.083	03:37
Air Blank		0.000	03:38
Subject Sample	# 1	0.180	03:38
Breath Volume		3.312	03:38
Air Blank		0.000	03:39
Air Blank		0.000	03:41
Subject Sample	# 2	0.179	03:41
Breath Volume		3.093	03:41
Air Blank		0.000	03:42
Control Test		0.083	03:42
Air Blank		0.000	03:43
Diagnostics Check		OK	03:43

Data Download Date/Time: 01/26/2016 04:01

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001112

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOLFF,ROSE,M
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: WOLFF,ROSE
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 12/28/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1686121
DRY GAS EXPIRY DATE: 05/21/2017
VIOLATION CODE: DUI
ARREST TIME: 22:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:17

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 01/23/2016 00:07:07

Diagnostics Check		OK	23:55	
Air Blank		0.000	23:56	
Control Test		0.082	23:56	
Air Blank		0.000	23:57	
Subject Sample	# 1		00:00	No Sample Provided
Breath Volume		0.000	00:00	
Air Blank		0.000	00:00	
Air Blank		0.000	00:02	
Subject Sample	# 2		00:05	No Sample Provided
Breath Volume		0.000	00:05	
Air Blank		0.000	00:06	
Control Test		0.082	00:06	
Air Blank		0.000	00:07	
Diagnostics Check		OK	00:07	

Data Download Date/Time: 01/26/2016 04:01

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001112

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOLFF,ROSE,M
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: WOLFF,ROSE
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 12/28/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1686121
DRY GAS EXPIRY DATE: 05/21/2017
VIOLATION CODE: DUI
ARREST TIME: 22:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:17

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 01/23/2016 00:23:57

Diagnostics Check		OK	00:12	
Air Blank		0.000	00:12	
Control Test		0.082	00:13	
Air Blank		0.000	00:13	
Subject Sample	# 1	0.116	00:16	Volume Not Met
Breath Volume		0.000	00:16	
Air Blank		0.000	00:17	
Air Blank		0.000	00:19	
Subject Sample	# 2	0.116	00:22	Volume Not Met
Breath Volume		0.093	00:22	
Air Blank		0.000	00:23	
Control Test		0.082	00:23	
Air Blank		0.000	00:23	
Diagnostics Check		OK	00:23	

Data Download Date/Time: 01/26/2016 04:01

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001112

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOLFF,ROSE,M
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: WOLFF,ROSE
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 12/28/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1686121
DRY GAS EXPIRY DATE: 05/21/2017
VIOLATION CODE: DUI
ARREST TIME: 22:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:17

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 01/23/2016 00:46:58

Diagnostics Check		OK	00:35	
Air Blank		0.000	00:35	
Control Test		0.082	00:36	
Air Blank		0.000	00:36	
Subject Sample	# 1	0.128	00:39	Volume Not Met
Breath Volume		0.261	00:39	
Air Blank		0.000	00:40	
Air Blank		0.000	00:42	
Subject Sample	# 2	0.136	00:45	Volume Not Met
Breath Volume		0.695	00:45	
Air Blank		0.000	00:46	
Control Test		0.080	00:46	
Air Blank		0.000	00:46	
Diagnostics Check		OK	00:46	

Data Download Date/Time: 01/26/2016 04:01

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001112

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HITCHCOCK,WILLIAM,D
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: HITCHCOCK,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 12/28/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1686121
DRY GAS EXPIRY DATE: 05/21/2017
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 01/24/2016 01:44:10

Diagnostics Check		OK	01:38
Air Blank		0.000	01:38
Control Test		0.084	01:38
Air Blank		0.000	01:39
Subject Sample	# 1	0.131	01:39
Breath Volume		3.390	01:39
Air Blank		0.000	01:40
Air Blank		0.000	01:42
Subject Sample	# 2	0.129	01:42
Breath Volume		2.789	01:42
Air Blank		0.000	01:43
Control Test		0.083	01:43
Air Blank		0.000	01:44
Diagnostics Check		OK	01:44

Data Download Date/Time: 01/26/2016 04:01

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001112

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MULLIGAN,WILLIAM,C
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: BEHRENSHOUSER,RICHARD
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 01/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1766638
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 12:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:50

SUBJECT'S LAST NAME: TRAWITZ

DATE/TIME OF TEST: 01/28/2016 14:28:07

Diagnostics Check		OK	14:20
Air Blank		0.000	14:20
Control Test		0.082	14:21
Air Blank		0.000	14:21
Subject Sample	# 1	0.000	14:23
Breath Volume		2.082	14:23
Air Blank		0.000	14:23
Air Blank		0.000	14:25
Subject Sample	# 2	0.000	14:26
Breath Volume		1.625	14:26
Air Blank		0.000	14:27
Control Test		0.082	14:27
Air Blank		0.000	14:28
Diagnostics Check		OK	14:28

Data Download Date/Time: 02/29/2016 00:38

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001114

INSTRUMENT REGISTERED TO: DELAND P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SCOTT,F
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: OWEN,SCOTT
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 01/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 133849
DRY GAS EXPIRY DATE: 10/12/2018
VIOLATION CODE: DUI
ARREST TIME: 20:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:46

SUBJECT'S LAST NAME: SOLIS

DATE/TIME OF TEST: 01/28/2016 21:54:11

Diagnostics Check		OK	21:47
Air Blank		0.000	21:47
Control Test		0.079	21:48
Air Blank		0.000	21:48
Subject Sample	# 1	0.205	21:49
Breath Volume		2.160	21:49
Air Blank		0.000	21:50
Air Blank		0.000	21:52
Subject Sample	# 2	0.217	21:52
Breath Volume		2.226	21:52
Air Blank		0.000	21:53
Control Test		0.078	21:53
Air Blank		0.000	21:54
Diagnostics Check		OK	21:54

Data Download Date/Time: 02/25/2016 18:53

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001116

INSTRUMENT REGISTERED TO: EGLIN AFB
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANDERS,JAMES,E
BREATH TEST OPERATOR AGENCY: Eglin AFB PD
ARRESTING OFFICER: PARRINO,TODD
ARRESTING OFFICER AGENCY: Eglin AFB PD
LAST AGENCY INSPECTION: 12/31/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26614080A7
DRY GAS EXPIRY DATE: 10/05/2016
VIOLATION CODE: DUI
ARREST TIME: 03:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: MERCER

DATE/TIME OF TEST: 01/17/2016 04:39:51

Diagnostics Check		OK	04:31
Air Blank		0.000	04:32
Control Test		0.080	04:32
Air Blank		0.000	04:33
Subject Sample	# 1	0.167	04:35
Breath Volume		2.652	04:35
Air Blank		0.000	04:35
Air Blank		0.000	04:37
Subject Sample	# 2	0.163	04:38
Breath Volume		3.757	04:38
Air Blank		0.000	04:38
Control Test		0.079	04:39
Air Blank		0.000	04:39
Diagnostics Check		OK	04:39

Data Download Date/Time: 02/02/2016 20:09

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001116

INSTRUMENT REGISTERED TO: EGLIN AFB
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRUNEAU,CINDY,L
BREATH TEST OPERATOR AGENCY: Eglin AFB PD
ARRESTING OFFICER: JOHNSON,MICHAEL
ARRESTING OFFICER AGENCY: Eglin AFB PD
LAST AGENCY INSPECTION: 12/31/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26614080A7
DRY GAS EXPIRY DATE: 10/05/2016
VIOLATION CODE: ZERO TOL
ARREST TIME: 09:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:43

SUBJECT'S LAST NAME: PRINCE

DATE/TIME OF TEST: 01/22/2016 11:14:39

Diagnostics Check		OK	11:04	
Air Blank		0.000	11:04	
Control Test		0.080	11:04	
Air Blank		0.000	11:05	
Subject Sample	# 1	0.095	11:08	Volume Not Met
Breath Volume		0.085	11:08	
Air Blank		0.000	11:09	
Air Blank		0.000	11:11	
Subject Sample	# 2		11:14	Sequence Aborted
Breath Volume		0.015	11:14	
Air Blank		0.000	11:14	

Data Download Date/Time: 02/02/2016 20:09

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001116

INSTRUMENT REGISTERED TO: EGLIN AFB
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRUNEAU,CINDY,L
BREATH TEST OPERATOR AGENCY: Eglin AFB PD
ARRESTING OFFICER: JOHNSON,MICHAEL
ARRESTING OFFICER AGENCY: Eglin AFB PD
LAST AGENCY INSPECTION: 12/31/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26614080A7
DRY GAS EXPIRY DATE: 10/05/2016
VIOLATION CODE: ZERO TOL
ARREST TIME: 09:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:43

SUBJECT'S LAST NAME: PRINCE

DATE/TIME OF TEST: 01/22/2016 11:19:49

Diagnostics Check		OK	11:16	
Air Blank		0.000	11:16	
Control Test		0.078	11:17	
Air Blank		0.000	11:17	
Subject Sample	# 1		11:19	Sequence Aborted
Breath Volume		0.015	11:19	
Air Blank		0.000	11:19	

Data Download Date/Time: 02/02/2016 20:09

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001116

INSTRUMENT REGISTERED TO: EGLIN AFB
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRUNEAU,CINDY,L
BREATH TEST OPERATOR AGENCY: Eglin AFB PD
ARRESTING OFFICER: BRUNEAU,CINDY
ARRESTING OFFICER AGENCY: Eglin AFB PD
LAST AGENCY INSPECTION: 12/31/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26614080A7
DRY GAS EXPIRY DATE: 10/05/2016
VIOLATION CODE: ADMIN
ARREST TIME: 09:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:43

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 01/22/2016 11:31:07

Diagnostics Check		OK	11:25
Air Blank		0.000	11:25
Control Test		0.080	11:25
Air Blank		0.000	11:26
Subject Sample	# 1	0.000	11:26
Breath Volume		2.492	11:26
Air Blank		0.000	11:27
Air Blank		0.000	11:29
Subject Sample	# 2	0.000	11:29
Breath Volume		2.312	11:29
Air Blank		0.000	11:30
Control Test		0.078	11:30
Air Blank		0.000	11:31
Diagnostics Check		OK	11:31

Data Download Date/Time: 02/02/2016 20:09

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001116

INSTRUMENT REGISTERED TO: EGLIN AFB
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRUNEAU,CINDY,L
BREATH TEST OPERATOR AGENCY: Eglin AFB PD
ARRESTING OFFICER: JOHNSON,MICHAEL
ARRESTING OFFICER AGENCY: Eglin AFB PD
LAST AGENCY INSPECTION: 12/31/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26614080A7
DRY GAS EXPIRY DATE: 10/05/2016
VIOLATION CODE: ZERO TOL
ARREST TIME: 09:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:43

SUBJECT'S LAST NAME: PRINCE

DATE/TIME OF TEST: 01/22/2016 11:39:39

Diagnostics Check		OK	11:32
Air Blank		0.000	11:33
Control Test		0.079	11:33
Air Blank		0.000	11:33
Subject Sample	# 1	0.081	11:34
Breath Volume		1.304	11:34
Air Blank		0.000	11:35
Air Blank		0.000	11:37
Subject Sample	# 2	0.082	11:38
Breath Volume		1.667	11:38
Air Blank		0.000	11:38
Control Test		0.079	11:39
Air Blank		0.000	11:39
Diagnostics Check		OK	11:39

Data Download Date/Time: 02/02/2016 20:09

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001116

INSTRUMENT REGISTERED TO: EGLIN AFB
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JAROSEWICZ,TRENTON,P
BREATH TEST OPERATOR AGENCY: Eglin AFB PD
ARRESTING OFFICER: FENNIG,CALEB
ARRESTING OFFICER AGENCY: Eglin AFB PD
LAST AGENCY INSPECTION: 12/31/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26614080A7
DRY GAS EXPIRY DATE: 10/05/2016
VIOLATION CODE: DUI
ARREST TIME: 03:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:23

SUBJECT'S LAST NAME: JAROSEWICZ

DATE/TIME OF TEST: 01/24/2016 04:56:27

Diagnostics Check		OK	04:48
Air Blank		0.000	04:48
Control Test		0.080	04:48
Air Blank		0.000	04:49
Subject Sample	# 1	0.138	04:51
Breath Volume		1.480	04:51
Air Blank		0.000	04:51
Air Blank		0.000	04:53
Subject Sample	# 2	0.129	04:54
Breath Volume		1.503	04:54
Air Blank		0.000	04:55
Control Test		0.080	04:55
Air Blank		0.000	04:56
Diagnostics Check		OK	04:56

Data Download Date/Time: 02/02/2016 20:09

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001116

INSTRUMENT REGISTERED TO: EGLIN AFB
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY,GARY,L
BREATH TEST OPERATOR AGENCY: Eglin AFB PD
ARRESTING OFFICER: BRISBANE,KADEEM
ARRESTING OFFICER AGENCY: Eglin AFB PD
LAST AGENCY INSPECTION: 12/31/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26614080A7
DRY GAS EXPIRY DATE: 10/05/2016
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: PRUETT

DATE/TIME OF TEST: 01/25/2016 02:26:29

Diagnostics Check		OK	02:19
Air Blank		0.000	02:20
Control Test		0.081	02:20
Air Blank		0.000	02:20
Subject Sample	# 1	0.253	02:21
Breath Volume		3.218	02:21
Air Blank		0.000	02:22
Air Blank		0.000	02:24
Subject Sample	# 2	0.253	02:24
Breath Volume		2.011	02:24
Air Blank		0.000	02:25
Control Test		0.079	02:25
Air Blank		0.000	02:26
Diagnostics Check		OK	02:26

Data Download Date/Time: 02/02/2016 20:09

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001119

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAWKINS,DEBORAH,
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: HAWKINS,DEBORAH
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 12/17/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28814080A3
DRY GAS EXPIRY DATE: 11/05/2016
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:56

SUBJECT'S LAST NAME: HANFORD

DATE/TIME OF TEST: 01/01/2016 03:29:20

Diagnostics Check		OK	03:21
Air Blank		0.000	03:22
Control Test		0.078	03:22
Air Blank		0.000	03:23
Subject Sample	# 1	0.117	03:24
Breath Volume		1.972	03:24
Air Blank		0.000	03:25
Air Blank		0.000	03:27
Subject Sample	# 2	0.117	03:27
Breath Volume		2.488	03:27
Air Blank		0.000	03:28
Control Test		0.077	03:28
Air Blank		0.000	03:29
Diagnostics Check		OK	03:29

Data Download Date/Time: 01/22/2016 19:44

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001119

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAWKINS,DEBORAH,
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: HAWKINS,DEBORAH
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 12/17/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28814080A3
DRY GAS EXPIRY DATE: 11/05/2016
VIOLATION CODE: DUI
ARREST TIME: 03:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:47

SUBJECT'S LAST NAME: GONZALEZ TORRES

DATE/TIME OF TEST: 01/03/2016 04:35:55

Diagnostics Check		OK	04:26	
Air Blank		0.000	04:27	
Control Test		0.081	04:27	
Air Blank		0.000	04:28	
Subject Sample	# 1		04:31	No Sample Provided
Breath Volume		0.000	04:31	
Air Blank		0.000	04:31	
Air Blank		0.000	04:33	
Subject Sample	# 2		04:34	Subject Test Refused
Breath Volume		0.000	04:34	
Air Blank		0.000	04:35	
Control Test		0.080	04:35	
Air Blank		0.000	04:35	
Diagnostics Check		OK	04:35	

Data Download Date/Time: 01/22/2016 19:44

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001121

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARIAS,ELOY,
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: ARIAS,ELOY
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 12/03/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1824329
DRY GAS EXPIRY DATE: 03/09/2018
VIOLATION CODE: DUI
ARREST TIME: 01:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: FRIEDMEN

DATE/TIME OF TEST: 01/01/2016 02:21:59

Diagnostics Check		OK	02:15
Air Blank		0.000	02:16
Control Test		0.081	02:16
Air Blank		0.000	02:16
Subject Sample	# 1	0.282	02:17
Breath Volume		2.265	02:17
Air Blank		0.000	02:18
Air Blank		0.000	02:19
Subject Sample	# 2	0.272	02:20
Breath Volume		1.730	02:20
Air Blank		0.000	02:21
Control Test		0.080	02:21
Air Blank		0.000	02:21
Diagnostics Check		OK	02:21

Data Download Date/Time: 01/22/2016 10:24

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001121

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANTANA,YUDIEL,
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: SANTANA,YUDIEL
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 12/03/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1824329
DRY GAS EXPIRY DATE: 03/09/2018
VIOLATION CODE: DUI
ARREST TIME: 03:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: SAYED

DATE/TIME OF TEST: 01/02/2016 04:24:13

Diagnostics Check		OK	04:20	
Air Blank		0.000	04:21	
Control Test		0.080	04:21	
Air Blank		0.000	04:22	
Subject Sample	# 1		04:22	Subject Test Refused
Breath Volume		0.000	04:22	
Air Blank		0.000	04:23	
Control Test		0.080	04:23	
Air Blank		0.000	04:24	
Diagnostics Check		OK	04:24	

Data Download Date/Time: 01/22/2016 10:24

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001121

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAHMING,DERRICK,A
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: RAHMING,DERRICK
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 12/03/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1824329
DRY GAS EXPIRY DATE: 03/09/2018
VIOLATION CODE: DUI
ARREST TIME: 22:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: BRODHEAD

DATE/TIME OF TEST: 01/05/2016 00:47:20

Diagnostics Check		OK	00:40
Air Blank		0.000	00:40
Control Test		0.082	00:40
Air Blank		0.000	00:41
Subject Sample	# 1	0.180	00:42
Breath Volume		2.835	00:42
Air Blank		0.000	00:43
Air Blank		0.000	00:45
Subject Sample	# 2	0.180	00:45
Breath Volume		2.492	00:45
Air Blank		0.000	00:46
Control Test		0.080	00:46
Air Blank		0.000	00:47
Diagnostics Check		OK	00:47

Data Download Date/Time: 01/22/2016 10:24

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001121

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEPESTRE,GREGORY,
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: DEPESTRE,GREGORY
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 12/03/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1824329
DRY GAS EXPIRY DATE: 03/09/2018
VIOLATION CODE: DUI
ARREST TIME: 23:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: RUIZ

DATE/TIME OF TEST: 01/06/2016 01:03:23

Diagnostics Check		OK	01:00	
Air Blank		0.000	01:00	
Control Test		0.082	01:01	
Air Blank		0.000	01:01	
Subject Sample	# 1		01:02	Subject Test Refused
Breath Volume		0.000	01:02	
Air Blank		0.000	01:02	
Control Test		0.081	01:02	
Air Blank		0.000	01:03	
Diagnostics Check		OK	01:03	

Data Download Date/Time: 01/22/2016 10:24

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001121

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARIAS,ELOY,
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: ARIAS,ELOY
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 12/03/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1824329
DRY GAS EXPIRY DATE: 03/09/2018
VIOLATION CODE: DUI
ARREST TIME: 01:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: JAUREGUI SOLIS

DATE/TIME OF TEST: 01/09/2016 02:00:41

Diagnostics Check		OK	01:53
Air Blank		0.000	01:54
Control Test		0.082	01:54
Air Blank		0.000	01:55
Subject Sample	# 1	0.151	01:55
Breath Volume		1.136	01:55
Air Blank		0.000	01:56
Air Blank		0.000	01:58
Subject Sample	# 2	0.148	01:59
Breath Volume		2.414	01:59
Air Blank		0.000	01:59
Control Test		0.081	02:00
Air Blank		0.000	02:00
Diagnostics Check		OK	02:00

Data Download Date/Time: 01/22/2016 10:24

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001121

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARTIGAS,MAYKOL,
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: ARTIGAS,MAYKOL
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 12/03/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1824329
DRY GAS EXPIRY DATE: 03/09/2018
VIOLATION CODE: DUI
ARREST TIME: 04:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:45

SUBJECT'S LAST NAME: VERA

DATE/TIME OF TEST: 01/09/2016 06:10:04

Diagnostics Check		OK	06:06	
Air Blank		0.000	06:06	
Control Test		0.082	06:06	
Air Blank		0.000	06:07	
Subject Sample	# 1		06:08	Subject Test Refused
Breath Volume		0.000	06:08	
Air Blank		0.000	06:09	
Control Test		0.081	06:09	
Air Blank		0.000	06:09	
Diagnostics Check		OK	06:10	

Data Download Date/Time: 01/22/2016 10:24

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001121

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAVINON,CARLOS,M
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: SAVINON,CARLOS
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 12/03/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1824329
DRY GAS EXPIRY DATE: 03/09/2018
VIOLATION CODE: DUI
ARREST TIME: 04:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:40

SUBJECT'S LAST NAME: KNOWLES

DATE/TIME OF TEST: 01/10/2016 05:04:12

Diagnostics Check		OK	05:01	
Air Blank		0.000	05:01	
Control Test		0.082	05:02	
Air Blank		0.000	05:02	
Subject Sample	# 1		05:02	Subject Test Refused
Breath Volume		0.000	05:02	
Air Blank		0.000	05:03	
Control Test		0.081	05:03	
Air Blank		0.000	05:04	
Diagnostics Check		OK	05:04	

Data Download Date/Time: 01/22/2016 10:24

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001121

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANTANA,YUDIEL,
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: SANTANA,YUDIEL
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 12/03/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1824329
DRY GAS EXPIRY DATE: 03/09/2018
VIOLATION CODE: DUI
ARREST TIME: 05:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:15

SUBJECT'S LAST NAME: PAEZ PELAEZ

DATE/TIME OF TEST: 01/10/2016 06:42:17

Diagnostics Check		OK	06:36
Air Blank		0.000	06:36
Control Test		0.082	06:36
Air Blank		0.000	06:37
Subject Sample	# 1	0.129	06:37
Breath Volume		1.324	06:37
Air Blank		0.000	06:38
Air Blank		0.000	06:40
Subject Sample	# 2	0.126	06:40
Breath Volume		2.121	06:40
Air Blank		0.000	06:41
Control Test		0.082	06:41
Air Blank		0.000	06:42
Diagnostics Check		OK	06:42

Data Download Date/Time: 01/22/2016 10:24

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001121

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARIAS,ELOY,
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: ARIAS,ELOY
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 12/03/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1824329
DRY GAS EXPIRY DATE: 03/09/2018
VIOLATION CODE: DUI
ARREST TIME: 04:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:30

SUBJECT'S LAST NAME: REED

DATE/TIME OF TEST: 01/16/2016 07:24:59

Diagnostics Check		OK	07:18
Air Blank		0.000	07:19
Control Test		0.080	07:19
Air Blank		0.000	07:19
Subject Sample	# 1	0.199	07:20
Breath Volume		3.382	07:20
Air Blank		0.000	07:21
Air Blank		0.000	07:22
Subject Sample	# 2	0.190	07:23
Breath Volume		3.472	07:23
Air Blank		0.000	07:24
Control Test		0.080	07:24
Air Blank		0.000	07:24
Diagnostics Check		OK	07:24

Data Download Date/Time: 01/22/2016 10:24

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001121

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HICKS,TERENCE,E
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: HICKS,TERENCE
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 12/03/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1824329
DRY GAS EXPIRY DATE: 03/09/2018
VIOLATION CODE: DUI
ARREST TIME: 04:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:40

SUBJECT'S LAST NAME: SAENZ

DATE/TIME OF TEST: 01/21/2016 05:07:18

Diagnostics Check		OK	05:01
Air Blank		0.000	05:01
Control Test		0.081	05:01
Air Blank		0.000	05:02
Subject Sample	# 1	0.148	05:02
Breath Volume		2.312	05:02
Air Blank		0.000	05:03
Air Blank		0.000	05:05
Subject Sample	# 2	0.147	05:05
Breath Volume		3.144	05:05
Air Blank		0.000	05:06
Control Test		0.082	05:06
Air Blank		0.000	05:07
Diagnostics Check		OK	05:07

Data Download Date/Time: 01/22/2016 10:24

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001121

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NANAN,KEIL,
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: NANAN,KEIL
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 01/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1824329
DRY GAS EXPIRY DATE: 03/09/2018
VIOLATION CODE: DUI
ARREST TIME: 20:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:20

SUBJECT'S LAST NAME: MENDOZA

DATE/TIME OF TEST: 01/22/2016 20:46:44

Diagnostics Check		OK	20:40
Air Blank		0.000	20:40
Control Test		0.081	20:41
Air Blank		0.000	20:41
Subject Sample	# 1	0.219	20:42
Breath Volume		1.769	20:42
Air Blank		0.000	20:42
Air Blank		0.000	20:44
Subject Sample	# 2	0.205	20:45
Breath Volume		1.121	20:45
Air Blank		0.000	20:45
Control Test		0.081	20:46
Air Blank		0.000	20:46
Diagnostics Check		OK	20:46

Data Download Date/Time: 02/27/2016 11:27

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001121

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEPESTRE,GREGORY,
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: DEPESTRE,GREGORY
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 01/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1824329
DRY GAS EXPIRY DATE: 03/09/2018
VIOLATION CODE: DUI
ARREST TIME: 04:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:07

SUBJECT'S LAST NAME: OQUENDO

DATE/TIME OF TEST: 01/23/2016 05:33:25

Diagnostics Check		OK	05:27
Air Blank		0.000	05:27
Control Test		0.082	05:28
Air Blank		0.000	05:28
Subject Sample	# 1	0.189	05:29
Breath Volume		2.398	05:29
Air Blank		0.000	05:29
Air Blank		0.000	05:31
Subject Sample	# 2	0.190	05:31
Breath Volume		2.480	05:31
Air Blank		0.000	05:32
Control Test		0.080	05:32
Air Blank		0.000	05:33
Diagnostics Check		OK	05:33

Data Download Date/Time: 02/27/2016 11:27

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001121

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROCA,LUIS,C
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: LLANIO,PIERRE
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 01/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1824329
DRY GAS EXPIRY DATE: 03/09/2018
VIOLATION CODE: DUI
ARREST TIME: 11:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 11:30

SUBJECT'S LAST NAME: LANZAS

DATE/TIME OF TEST: 01/28/2016 11:59:56

Diagnostics Check		OK	11:52
Air Blank		0.000	11:52
Control Test		0.082	11:52
Air Blank		0.000	11:53
Subject Sample	# 1	0.099	11:55
Breath Volume		1.257	11:55
Air Blank		0.000	11:56
Air Blank		0.000	11:58
Subject Sample	# 2	0.099	11:58
Breath Volume		2.351	11:58
Air Blank		0.000	11:59
Control Test		0.079	11:59
Air Blank		0.000	11:59
Diagnostics Check		OK	11:59

Data Download Date/Time: 02/27/2016 11:27

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001121

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MATTHEWS,LAURIE,A
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: PEREZ,NELSON
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 01/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1824329
DRY GAS EXPIRY DATE: 03/09/2018
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:53

SUBJECT'S LAST NAME: MATTHEWS

DATE/TIME OF TEST: 01/30/2016 03:18:59

Diagnostics Check		OK	03:16	
Air Blank		0.000	03:16	
Control Test		0.082	03:16	
Air Blank		0.000	03:17	
Subject Sample	# 1		03:17	Subject Test Refused
Breath Volume		0.000	03:17	
Air Blank		0.000	03:18	
Control Test		0.082	03:18	
Air Blank		0.000	03:18	
Diagnostics Check		OK	03:18	

Data Download Date/Time: 02/27/2016 11:28

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001121

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,NELSON,
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: PEREZ,NELSON
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 01/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1824329
DRY GAS EXPIRY DATE: 03/09/2018
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:53

SUBJECT'S LAST NAME: MATTHEWS

DATE/TIME OF TEST: 01/30/2016 03:26:19

Diagnostics Check		OK	03:23	
Air Blank		0.000	03:23	
Control Test		0.082	03:24	
Air Blank		0.000	03:24	
Subject Sample	# 1		03:24	Subject Test Refused
Breath Volume		0.000	03:24	
Air Blank		0.000	03:25	
Control Test		0.082	03:25	
Air Blank		0.000	03:26	
Diagnostics Check		OK	03:26	

Data Download Date/Time: 02/27/2016 11:28

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001121

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALLO,DENNIS,
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: GALLO,DENNIS
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 01/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1824329
DRY GAS EXPIRY DATE: 03/09/2018
VIOLATION CODE: DUI
ARREST TIME: 03:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:13

SUBJECT'S LAST NAME: CALIX GARCIA

DATE/TIME OF TEST: 01/31/2016 04:51:28

Diagnostics Check		OK	04:45	
Air Blank		0.000	04:45	
Control Test		0.080	04:46	
Air Blank		0.000	04:46	
Subject Sample	# 1		04:49	No Sample Provided
Breath Volume		0.000	04:49	
Air Blank			04:50	RFI Detect
Air Blank		0.000	04:50	
Control Test		0.082	04:50	
Air Blank		0.000	04:51	
Diagnostics Check		OK	04:51	

Data Download Date/Time: 02/27/2016 11:28

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001121

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALLO,DENNIS,
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: GALLO,DENNIS
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 01/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1824329
DRY GAS EXPIRY DATE: 03/09/2018
VIOLATION CODE: DUI
ARREST TIME: 03:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:13

SUBJECT'S LAST NAME: CALIX GARCIA

DATE/TIME OF TEST: 01/31/2016 04:59:42

Diagnostics Check		OK	04:55	
Air Blank		0.000	04:55	
Control Test		0.081	04:55	
Air Blank		0.000	04:56	
Subject Sample	# 1		04:58	Subject Test Refused
Breath Volume		0.000	04:58	
Air Blank		0.000	04:58	
Control Test		0.082	04:59	
Air Blank		0.000	04:59	
Diagnostics Check		OK	04:59	

Data Download Date/Time: 02/27/2016 11:28

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001121

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALLO,DENNIS,
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: GALLO,DENNIS
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 01/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1824329
DRY GAS EXPIRY DATE: 03/09/2018
VIOLATION CODE: DUI
ARREST TIME: 03:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:13

SUBJECT'S LAST NAME: CALIX GARCIA

DATE/TIME OF TEST: 01/31/2016 05:14:15

Diagnostics Check		OK	05:08	
Air Blank		0.000	05:09	
Control Test		0.082	05:09	
Air Blank		0.000	05:10	
Subject Sample	# 1		05:12	Subject Test Refused
Breath Volume		0.070	05:12	
Air Blank		0.000	05:13	
Control Test		0.081	05:13	
Air Blank		0.000	05:14	
Diagnostics Check		OK	05:14	

Data Download Date/Time: 02/27/2016 11:28

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001121

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALLO,DENNIS,
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: GALLO,DENNIS
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 01/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1824329
DRY GAS EXPIRY DATE: 03/09/2018
VIOLATION CODE: DUI
ARREST TIME: 03:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:13

SUBJECT'S LAST NAME: CALIX GARCIA

DATE/TIME OF TEST: 01/31/2016 05:24:30

Diagnostics Check		OK	05:16
Air Blank		0.000	05:16
Control Test		0.080	05:17
Air Blank		0.000	05:17
Subject Sample	# 1	0.172	05:20
Breath Volume		1.386	05:20
Air Blank		0.000	05:20
Air Blank		0.000	05:22
Subject Sample	# 2	0.186	05:22
Breath Volume		3.074	05:22
Air Blank		0.000	05:23
Control Test		0.077	05:23
Air Blank		0.000	05:24
Diagnostics Check		OK	05:24

Data Download Date/Time: 02/27/2016 11:28

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001128

INSTRUMENT REGISTERED TO: PUTNAM COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BEDENBAUGH,JUSTIN,H
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: BEDENBAUGH,JUSTIN
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 12/31/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 3131108A1
DRY GAS EXPIRY DATE: 04/05/2017
VIOLATION CODE: DUI
ARREST TIME: 06:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:28

SUBJECT'S LAST NAME: WAHL

DATE/TIME OF TEST: 01/04/2016 08:07:36

Diagnostics Check		OK	08:01
Air Blank		0.000	08:01
Control Test		0.079	08:01
Air Blank		0.000	08:02
Subject Sample	# 1	0.178	08:03
Breath Volume		1.992	08:03
Air Blank		0.000	08:03
Air Blank		0.000	08:05
Subject Sample	# 2	0.186	08:06
Breath Volume		1.554	08:06
Air Blank		0.000	08:06
Control Test		0.078	08:07
Air Blank		0.000	08:07
Diagnostics Check		OK	08:07

Data Download Date/Time: 01/31/2016 01:59

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001128

INSTRUMENT REGISTERED TO: PUTNAM COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRENCH,CALEB,L
BREATH TEST OPERATOR AGENCY: Putnam County SO
ARRESTING OFFICER: JOWERS,THOMAS
ARRESTING OFFICER AGENCY: Putnam County SO
LAST AGENCY INSPECTION: 12/31/2015

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 07515080A4
DRY GAS EXPIRY DATE: 04/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:16

SUBJECT'S LAST NAME: ROSADO

DATE/TIME OF TEST: 01/08/2016 21:12:53

Diagnostics Check		OK	21:03	
Air Blank		0.000	21:03	
Control Test		0.078	21:04	
Air Blank		0.000	21:04	
Subject Sample	# 1	0.132	21:07	Volume Not Met
Breath Volume		0.406	21:07	
Air Blank		0.000	21:08	
Air Blank		0.000	21:10	
Subject Sample	# 2		21:11	Subject Test Refused
Breath Volume		0.066	21:11	
Air Blank		0.000	21:11	
Control Test		0.077	21:12	
Air Blank		0.000	21:12	
Diagnostics Check		OK	21:12	

Data Download Date/Time: 01/31/2016 01:59

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001128

INSTRUMENT REGISTERED TO: PUTNAM COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,TIFFANY,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: WILLIAMS,TIFFANY
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 12/31/2015

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07515080A4
DRY GAS EXPIRY DATE: 04/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: MORTON

DATE/TIME OF TEST: 01/30/2016 03:17:27

Diagnostics Check		OK	03:10
Air Blank		0.000	03:11
Control Test		0.079	03:11
Air Blank		0.000	03:12
Subject Sample	# 1	0.118	03:12
Breath Volume		1.972	03:12
Air Blank		0.000	03:13
Air Blank		0.000	03:15
Subject Sample	# 2	0.125	03:15
Breath Volume		2.484	03:15
Air Blank		0.000	03:16
Control Test		0.079	03:16
Air Blank		0.000	03:17
Diagnostics Check		OK	03:17

Data Download Date/Time: 01/31/2016 01:59

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001128

INSTRUMENT REGISTERED TO: PUTNAM COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,TIFFANY,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: WILLIAMS,TIFFANY
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 01/31/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07515080A4
DRY GAS EXPIRY DATE: 04/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:13

SUBJECT'S LAST NAME: RUBIO

DATE/TIME OF TEST: 01/31/2016 03:41:26

Diagnostics Check		OK	03:34
Air Blank		0.000	03:35
Control Test		0.079	03:35
Air Blank		0.000	03:36
Subject Sample	# 1	0.209	03:36
Breath Volume		1.507	03:36
Air Blank		0.000	03:37
Air Blank		0.000	03:39
Subject Sample	# 2	0.208	03:39
Breath Volume		1.414	03:39
Air Blank		0.000	03:40
Control Test		0.079	03:40
Air Blank		0.000	03:41
Diagnostics Check		OK	03:41

Data Download Date/Time: 03/03/2016 06:11

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001130

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PECK,KEITH,D
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: PECK,KEITH
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 01/20/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1589901
DRY GAS EXPIRY DATE: 10/30/2016
VIOLATION CODE: DUI
ARREST TIME: 21:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:14

SUBJECT'S LAST NAME: LEAGLE

DATE/TIME OF TEST: 01/30/2016 22:41:12

Diagnostics Check		OK	22:34
Air Blank		0.000	22:35
Control Test		0.079	22:35
Air Blank		0.000	22:36
Subject Sample	# 1	0.145	22:36
Breath Volume		1.800	22:36
Air Blank		0.000	22:37
Air Blank		0.000	22:39
Subject Sample	# 2	0.152	22:39
Breath Volume		1.980	22:39
Air Blank		0.000	22:40
Control Test		0.078	22:40
Air Blank		0.000	22:41
Diagnostics Check		OK	22:41

Data Download Date/Time: 02/17/2016 15:04

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001130

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PECK,KEITH,D
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: FROST,WES
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 01/20/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1589901
DRY GAS EXPIRY DATE: 10/30/2016
VIOLATION CODE: DUI
ARREST TIME: 01:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:08

SUBJECT'S LAST NAME: COLEMAN

DATE/TIME OF TEST: 01/31/2016 02:36:51

Diagnostics Check		OK	02:29
Air Blank		0.000	02:29
Control Test		0.078	02:30
Air Blank		0.000	02:30
Subject Sample	# 1	0.179	02:31
Breath Volume		2.003	02:31
Air Blank		0.000	02:32
Air Blank		0.000	02:33
Subject Sample	# 2	0.174	02:35
Breath Volume		2.136	02:35
Air Blank		0.000	02:35
Control Test		0.078	02:36
Air Blank		0.000	02:36
Diagnostics Check		OK	02:36

Data Download Date/Time: 02/17/2016 15:04

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.