

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINA,EDDYMAR,Z
BREATH TEST OPERATOR AGENCY: Hialeah Gardens PD
ARRESTING OFFICER: MEDINA,EDDYMAR
ARRESTING OFFICER AGENCY: Hialeah Gardens PD
LAST AGENCY INSPECTION: 02/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25413080A2
DRY GAS EXPIRY DATE: 10/01/2015
VIOLATION CODE: DUI
ARREST TIME: 20:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:40

SUBJECT'S LAST NAME: MIGOYO

DATE/TIME OF TEST: 03/17/2016 21:07:37

Diagnostics Check	OK	21:03	
Air Blank	0.000	21:04	
Control Test	0.076	21:04	
Air Blank	0.000	21:04	
Subject Sample	# 1	21:06	Subject Test Refused
Breath Volume	0.000	21:06	
Air Blank	0.000	21:06	
Control Test	0.078	21:06	
Air Blank	0.000	21:07	
Diagnostics Check	OK	21:07	

Data Download Date/Time: 03/31/2016 12:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORENO, RICARDO, J
BREATH TEST OPERATOR AGENCY: Aventura PD
ARRESTING OFFICER: BURNS, JEFFREY
ARRESTING OFFICER AGENCY: Aventura PD
LAST AGENCY INSPECTION: 02/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 00:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:16

SUBJECT'S LAST NAME: QUILES

DATE/TIME OF TEST: 03/09/2016 01:37:50

Diagnostics Check	OK	01:36	
Air Blank	0.000	01:36	
Control Test	0.072	01:37	Control Outside Tolerance
Air Blank	0.000	01:37	

Data Download Date/Time: 03/23/2016 10:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORENO, RICARDO, J
BREATH TEST OPERATOR AGENCY: Aventura PD
ARRESTING OFFICER: BURNS, JEFFREY
ARRESTING OFFICER AGENCY: Aventura PD
LAST AGENCY INSPECTION: 02/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 00:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:16

SUBJECT'S LAST NAME: QUILES SUBIRANA

DATE/TIME OF TEST: 03/09/2016 01:44:41

Diagnostics Check	OK	01:39	
Air Blank	0.000	01:39	
Control Test	0.079	01:40	
Air Blank	0.000	01:40	
Subject Sample	# 1 0.248	01:43	
Breath Volume	1.648	01:43	
Air Blank		01:43	Purge Fail
Air Blank	0.000	01:44	

Data Download Date/Time: 03/23/2016 10:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORENO,RICARDO,J
BREATH TEST OPERATOR AGENCY: Aventura PD
ARRESTING OFFICER: BURNS,JEFFREY
ARRESTING OFFICER AGENCY: Aventura PD
LAST AGENCY INSPECTION: 02/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 00:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:16

SUBJECT'S LAST NAME: QUILES SUBIRANA

DATE/TIME OF TEST: 03/09/2016 01:54:39

Diagnostics Check	OK	01:46
Air Blank	0.000	01:46
Control Test	0.076	01:46
Air Blank	0.000	01:47
Subject Sample # 1	0.279	01:49
Breath Volume	2.214	01:49
Air Blank	0.000	01:50
Air Blank	0.000	01:52
Subject Sample # 2	0.273	01:53
Breath Volume	2.613	01:53
Air Blank	0.000	01:53
Control Test	0.076	01:54
Air Blank	0.000	01:54
Diagnostics Check	OK	01:54

Data Download Date/Time: 03/23/2016 10:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHUMACHER,ERIC,L
BREATH TEST OPERATOR AGENCY: Aventura PD
ARRESTING OFFICER: BURNS,JEFFREY
ARRESTING OFFICER AGENCY: Aventura PD
LAST AGENCY INSPECTION: 02/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 16:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:25

SUBJECT'S LAST NAME: SHUMACHER

DATE/TIME OF TEST: 03/12/2016 17:02:38

Diagnostics Check	OK	16:56
Air Blank	0.000	16:56
Control Test	0.078	16:56
Air Blank	0.000	16:57
Subject Sample # 1	0.210	16:58
Breath Volume	2.984	16:58
Air Blank	0.000	16:58
Air Blank	0.000	17:00
Subject Sample # 2	0.218	17:01
Breath Volume	2.851	17:01
Air Blank	0.000	17:01
Control Test	0.076	17:02
Air Blank	0.000	17:02
Diagnostics Check	OK	17:02

Data Download Date/Time: 03/23/2016 10:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCGRATH,ROBERT,E
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: MCGRATH,ROBERT
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 02/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 08:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:02

SUBJECT'S LAST NAME: MIRANDA

DATE/TIME OF TEST: 03/13/2016 09:30:03

Diagnostics Check	OK	09:22
Air Blank	0.000	09:23
Control Test	0.079	09:23
Air Blank	0.000	09:24
Subject Sample # 1	0.140	09:25
Breath Volume	2.160	09:25
Air Blank	0.000	09:25
Air Blank	0.000	09:27
Subject Sample # 2	0.142	09:28
Breath Volume	2.171	09:28
Air Blank	0.000	09:29
Control Test	0.079	09:29
Air Blank	0.000	09:29
Diagnostics Check	OK	09:30

Data Download Date/Time: 03/23/2016 10:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WENDLING,JOSHUA,A
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: WENDLING,JOSHUA
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 02/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 17:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:51

SUBJECT'S LAST NAME: YAKUBOV

DATE/TIME OF TEST: 03/13/2016 18:24:31

Diagnostics Check	OK	18:21	
Air Blank	0.000	18:21	
Control Test	0.080	18:22	
Air Blank	0.000	18:22	
Subject Sample	# 1	18:23	Subject Test Refused
Breath Volume	0.000	18:23	
Air Blank	0.000	18:23	
Control Test	0.078	18:23	
Air Blank	0.000	18:24	
Diagnostics Check	OK	18:24	

Data Download Date/Time: 03/23/2016 10:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AGINS,LYLE,P
BREATH TEST OPERATOR AGENCY: Sunny Isles Beach PD
ARRESTING OFFICER: RUIZ,L
ARRESTING OFFICER AGENCY: Surfside PD
LAST AGENCY INSPECTION: 02/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 02:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:21

SUBJECT'S LAST NAME: SAKERS

DATE/TIME OF TEST: 03/14/2016 02:50:20

Diagnostics Check	OK	02:47	
Air Blank	0.000	02:47	
Control Test	0.080	02:48	
Air Blank	0.000	02:48	
Subject Sample	# 1	02:48	Subject Test Refused
Breath Volume	0.000	02:48	
Air Blank	0.000	02:49	
Control Test	0.079	02:49	
Air Blank	0.000	02:50	
Diagnostics Check	OK	02:50	

Data Download Date/Time: 03/23/2016 10:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAESTRE,HANS,
BREATH TEST OPERATOR AGENCY: Aventura PD
ARRESTING OFFICER: CRAIG,JOSEPH
ARRESTING OFFICER AGENCY: Aventura PD
LAST AGENCY INSPECTION: 02/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 19:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:22

SUBJECT'S LAST NAME: WALTERS

DATE/TIME OF TEST: 03/16/2016 20:21:11

Diagnostics Check		OK	20:14
Air Blank		0.000	20:14
Control Test		0.079	20:15
Air Blank		0.000	20:15
Subject Sample	# 1	0.156	20:16
Breath Volume		1.953	20:16
Air Blank		0.000	20:17
Air Blank		0.000	20:19
Subject Sample	# 2	0.163	20:19
Breath Volume		2.273	20:19
Air Blank		0.000	20:20
Control Test		0.077	20:20
Air Blank		0.000	20:21
Diagnostics Check		OK	20:21

Data Download Date/Time: 03/23/2016 10:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUDAT,CLAUDIA,I
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: ANDUJA,ANNELLIS
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 02/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 18:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:27

SUBJECT'S LAST NAME: DOLCE

DATE/TIME OF TEST: 03/20/2016 19:22:23

Diagnostics Check	OK	19:19	
Air Blank	0.000	19:19	
Control Test	0.079	19:20	
Air Blank	0.000	19:20	
Subject Sample	# 1	19:21	Subject Test Refused
Breath Volume	0.000	19:21	
Air Blank	0.000	19:21	
Control Test	0.079	19:21	
Air Blank	0.000	19:22	
Diagnostics Check	OK	19:22	

Data Download Date/Time: 03/23/2016 10:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: KERSHAW,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/09/2017
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:01

SUBJECT'S LAST NAME: BULAS

DATE/TIME OF TEST: 03/05/2016 04:24:52

Diagnostics Check	OK	04:22	
Air Blank	0.000	04:22	
Control Test	0.082	04:22	
Air Blank	0.000	04:23	
Subject Sample	# 1	04:23	Subject Test Refused
Breath Volume	0.000	04:23	
Air Blank	0.000	04:23	
Control Test	0.081	04:24	
Air Blank	0.000	04:24	
Diagnostics Check	OK	04:24	

Data Download Date/Time: 03/15/2016 12:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FISETTE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/09/2017
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:32

SUBJECT'S LAST NAME: LAWLER

DATE/TIME OF TEST: 03/08/2016 02:55:30

Diagnostics Check	OK	02:52	
Air Blank	0.000	02:53	
Control Test	0.082	02:53	
Air Blank	0.000	02:54	
Subject Sample	# 1	02:54	Subject Test Refused
Breath Volume	0.000	02:54	
Air Blank	0.000	02:54	
Control Test	0.081	02:54	
Air Blank	0.000	02:55	
Diagnostics Check	OK	02:55	

Data Download Date/Time: 03/15/2016 12:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIALKOSKI,WILLIAM,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: REAGOR,C.E
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/09/2017
VIOLATION CODE: DUI
ARREST TIME: 15:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:57

SUBJECT'S LAST NAME: DAUGHERTY

DATE/TIME OF TEST: 03/08/2016 17:32:24

Diagnostics Check		OK	17:18	
Air Blank		0.000	17:18	
Control Test		0.081	17:19	
Air Blank		0.000	17:19	
Subject Sample	# 1	0.157	17:22	Volume Not Met
Breath Volume		0.246	17:22	
Air Blank		0.000	17:23	
Air Blank		0.000	17:25	
Subject Sample	# 2	0.220	17:26	
Breath Volume		1.125	17:26	
Air Blank		0.000	17:27	
Air Blank		0.000	17:28	
Subject Sample	# 3	0.204	17:30	
Breath Volume		2.378	17:30	
Air Blank		0.000	17:31	
Control Test		0.080	17:31	
Air Blank		0.000	17:32	
Diagnostics Check		OK	17:32	

Data Download Date/Time: 03/15/2016 12:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIALKOSKI,WILLIAM,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: TERRELL,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/09/2017
VIOLATION CODE: DUI
ARREST TIME: 16:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:25

SUBJECT'S LAST NAME: HALL

DATE/TIME OF TEST: 03/08/2016 18:48:50

Diagnostics Check	OK	18:45	
Air Blank	0.000	18:46	
Control Test	0.080	18:46	
Air Blank	0.000	18:47	
Subject Sample	# 1	18:47	Subject Test Refused
Breath Volume	0.000	18:47	
Air Blank	0.000	18:47	
Control Test	0.080	18:48	
Air Blank	0.000	18:48	
Diagnostics Check	OK	18:48	

Data Download Date/Time: 03/15/2016 12:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIALKOSKI,WILLIAM,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FISETTE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/09/2017
VIOLATION CODE: DUI
ARREST TIME: 19:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:32

SUBJECT'S LAST NAME: HEMINGWAY

DATE/TIME OF TEST: 03/08/2016 20:58:20

Diagnostics Check	OK	20:52
Air Blank	0.000	20:52
Control Test	0.081	20:52
Air Blank	0.000	20:53
Subject Sample # 1	0.000	20:54
Breath Volume	2.875	20:54
Air Blank	0.000	20:54
Air Blank	0.000	20:56
Subject Sample # 2	0.000	20:56
Breath Volume	2.824	20:56
Air Blank	0.000	20:57
Control Test	0.081	20:57
Air Blank	0.000	20:58
Diagnostics Check	OK	20:58

Data Download Date/Time: 03/15/2016 12:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIALKOSKI,WILLIAM,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BARTILUCCI,D
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/09/2017
VIOLATION CODE: DUI
ARREST TIME: 20:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:29

SUBJECT'S LAST NAME: SPENCER

DATE/TIME OF TEST: 03/08/2016 21:56:19

Diagnostics Check	OK	21:50
Air Blank	0.000	21:50
Control Test	0.081	21:50
Air Blank	0.000	21:51
Subject Sample # 1	0.131	21:51
Breath Volume	1.320	21:51
Air Blank	0.000	21:52
Air Blank	0.000	21:54
Subject Sample # 2	0.141	21:54
Breath Volume	1.660	21:54
Air Blank	0.000	21:55
Control Test	0.080	21:55
Air Blank	0.000	21:56
Diagnostics Check	OK	21:56

Data Download Date/Time: 03/15/2016 12:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHITE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/09/2017
VIOLATION CODE: DUI
ARREST TIME: 21:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:14

SUBJECT'S LAST NAME: MCCOLLUM

DATE/TIME OF TEST: 03/08/2016 23:40:47

Diagnostics Check	OK	23:34
Air Blank	0.000	23:34
Control Test	0.081	23:35
Air Blank	0.000	23:35
Subject Sample # 1	0.000	23:36
Breath Volume	1.164	23:36
Air Blank	0.000	23:36
Air Blank	0.000	23:38
Subject Sample # 2	0.000	23:39
Breath Volume	1.574	23:39
Air Blank	0.000	23:39
Control Test	0.082	23:40
Air Blank	0.000	23:40
Diagnostics Check	OK	23:40

Data Download Date/Time: 03/15/2016 12:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HOOVER,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/09/2017
VIOLATION CODE: DUI
ARREST TIME: 01:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:37

SUBJECT'S LAST NAME: CARROLL

DATE/TIME OF TEST: 03/09/2016 03:02:16

Diagnostics Check	OK	02:59	
Air Blank	0.000	02:59	
Control Test	0.081	03:00	
Air Blank	0.000	03:00	
Subject Sample	# 1	03:00	Subject Test Refused
Breath Volume	0.000	03:00	
Air Blank	0.000	03:01	
Control Test	0.081	03:01	
Air Blank	0.000	03:02	
Diagnostics Check	OK	03:02	

Data Download Date/Time: 03/15/2016 12:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: TERRELL,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/09/2017
VIOLATION CODE: DUI
ARREST TIME: 22:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:13

SUBJECT'S LAST NAME: POSTOAK

DATE/TIME OF TEST: 03/09/2016 23:40:13

Diagnostics Check	OK	23:33
Air Blank	0.000	23:34
Control Test	0.080	23:34
Air Blank	0.000	23:34
Subject Sample # 1	0.109	23:35
Breath Volume	2.656	23:35
Air Blank	0.000	23:36
Air Blank	0.000	23:38
Subject Sample # 2	0.108	23:38
Breath Volume	3.023	23:38
Air Blank	0.000	23:39
Control Test	0.081	23:39
Air Blank	0.000	23:40
Diagnostics Check	OK	23:40

Data Download Date/Time: 03/15/2016 12:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HOOVER,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/09/2017
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: EUBANKS-BURKHEAD

DATE/TIME OF TEST: 03/10/2016 03:31:20

Diagnostics Check	OK	03:25
Air Blank	0.000	03:25
Control Test	0.081	03:25
Air Blank	0.000	03:26
Subject Sample # 1	0.056	03:26
Breath Volume	3.792	03:26
Air Blank	0.000	03:27
Air Blank	0.000	03:29
Subject Sample # 2	0.055	03:29
Breath Volume	3.597	03:29
Air Blank	0.000	03:30
Control Test	0.081	03:30
Air Blank	0.000	03:31
Diagnostics Check	OK	03:31

Data Download Date/Time: 03/15/2016 12:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BALS
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/09/2017
VIOLATION CODE: DUI
ARREST TIME: 22:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:41

SUBJECT'S LAST NAME: THIGPEN

DATE/TIME OF TEST: 03/12/2016 00:09:06

Diagnostics Check		OK	00:01
Air Blank		0.000	00:02
Control Test		0.081	00:02
Air Blank		0.000	00:03
Subject Sample	# 1	0.053	00:04
Breath Volume		2.082	00:04
Air Blank		0.000	00:05
Air Blank		0.000	00:06
Subject Sample	# 2	0.051	00:07
Breath Volume		2.375	00:07
Air Blank		0.000	00:08
Control Test		0.081	00:08
Air Blank		0.000	00:09
Diagnostics Check		OK	00:09

Data Download Date/Time: 03/15/2016 12:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: EANES,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/09/2017
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:37

SUBJECT'S LAST NAME: DEMONT

DATE/TIME OF TEST: 03/12/2016 02:59:57

Diagnostics Check	OK	02:57	
Air Blank	0.000	02:57	
Control Test	0.081	02:57	
Air Blank	0.000	02:58	
Subject Sample	# 1	02:58	Subject Test Refused
Breath Volume	0.000	02:58	
Air Blank	0.000	02:59	
Control Test	0.080	02:59	
Air Blank	0.000	02:59	
Diagnostics Check	OK	02:59	

Data Download Date/Time: 03/15/2016 12:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: COLEMAN,F
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/09/2017
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:41

SUBJECT'S LAST NAME: SYMITCZK

DATE/TIME OF TEST: 03/13/2016 03:07:38

Diagnostics Check	OK	03:01
Air Blank	0.000	03:02
Control Test	0.081	03:02
Air Blank	0.000	03:02
Subject Sample # 1	0.184	03:03
Breath Volume	1.726	03:03
Air Blank	0.000	03:03
Air Blank	0.000	03:05
Subject Sample # 2	0.188	03:06
Breath Volume	2.148	03:06
Air Blank	0.000	03:06
Control Test	0.080	03:07
Air Blank	0.000	03:07
Diagnostics Check	OK	03:07

Data Download Date/Time: 03/15/2016 12:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: KERSHAW,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/09/2017
VIOLATION CODE: DUI
ARREST TIME: 03:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:38

SUBJECT'S LAST NAME: FRANCEESCONI

DATE/TIME OF TEST: 03/13/2016 05:13:09

Diagnostics Check		OK	04:59	
Air Blank		0.000	04:59	
Control Test		0.081	04:59	
Air Blank		0.000	05:00	
Subject Sample	# 1	0.120	05:03	Volume Not Met
Breath Volume		0.664	05:03	
Air Blank		0.000	05:04	
Air Blank		0.000	05:06	
Subject Sample	# 2	0.160	05:07	
Breath Volume		1.203	05:07	
Air Blank		0.000	05:08	
Air Blank		0.000	05:10	
Subject Sample	# 3	0.157	05:11	
Breath Volume		1.347	05:11	
Air Blank		0.000	05:12	
Control Test		0.079	05:12	
Air Blank		0.000	05:13	
Diagnostics Check		OK	05:13	

Data Download Date/Time: 03/15/2016 12:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: EANES,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/09/2017
VIOLATION CODE: DUI
ARREST TIME: 23:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:48

SUBJECT'S LAST NAME: DUPRE

DATE/TIME OF TEST: 03/14/2016 01:11:26

Diagnostics Check	OK	01:08	
Air Blank	0.000	01:09	
Control Test	0.081	01:09	
Air Blank	0.000	01:09	
Subject Sample	# 1	01:10	Subject Test Refused
Breath Volume	0.000	01:10	
Air Blank	0.000	01:10	
Control Test	0.080	01:10	
Air Blank	0.000	01:11	
Diagnostics Check	OK	01:11	

Data Download Date/Time: 03/15/2016 12:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MONTGOMERY,K
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 02/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/09/2017
VIOLATION CODE: DUI
ARREST TIME: 02:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:41

SUBJECT'S LAST NAME: CERCY

DATE/TIME OF TEST: 03/15/2016 05:05:43

Diagnostics Check	OK	05:02	
Air Blank	0.000	05:03	
Control Test	0.080	05:03	
Air Blank	0.000	05:04	
Subject Sample	# 1	05:04	Subject Test Refused
Breath Volume	0.000	05:04	
Air Blank	0.000	05:04	
Control Test	0.081	05:05	
Air Blank	0.000	05:05	
Diagnostics Check	OK	05:05	

Data Download Date/Time: 03/15/2016 12:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: TERRELL,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/09/2017
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:09

SUBJECT'S LAST NAME: MOORE

DATE/TIME OF TEST: 03/16/2016 02:36:41

Diagnostics Check	OK	02:29
Air Blank	0.000	02:30
Control Test	0.081	02:30
Air Blank	0.000	02:31
Subject Sample # 1	0.119	02:32
Breath Volume	2.109	02:32
Air Blank	0.000	02:32
Air Blank	0.000	02:34
Subject Sample # 2	0.128	02:35
Breath Volume	1.253	02:35
Air Blank	0.000	02:35
Control Test	0.080	02:36
Air Blank	0.000	02:36
Diagnostics Check	OK	02:36

Data Download Date/Time: 03/31/2016 12:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MONTGOMERY,K
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/09/2017
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:49

SUBJECT'S LAST NAME: IRVIN

DATE/TIME OF TEST: 03/16/2016 03:15:34

Diagnostics Check	OK	03:09
Air Blank	0.000	03:10
Control Test	0.079	03:10
Air Blank	0.000	03:10
Subject Sample # 1	0.150	03:11
Breath Volume	2.859	03:11
Air Blank	0.000	03:11
Air Blank	0.000	03:13
Subject Sample # 2	0.147	03:14
Breath Volume	2.218	03:14
Air Blank	0.000	03:14
Control Test	0.080	03:14
Air Blank	0.000	03:15
Diagnostics Check	OK	03:15

Data Download Date/Time: 03/31/2016 12:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: TERRELL,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/09/2017
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:58

SUBJECT'S LAST NAME: KENNEDY

DATE/TIME OF TEST: 03/17/2016 01:23:56

Diagnostics Check	OK	01:18	
Air Blank	0.000	01:18	
Control Test	0.081	01:19	
Air Blank	0.000	01:19	
Subject Sample	# 1	01:22	Subject Test Refused
Breath Volume	0.000	01:22	
Air Blank	0.000	01:23	
Control Test	0.081	01:23	
Air Blank	0.000	01:23	
Diagnostics Check	OK	01:23	

Data Download Date/Time: 03/31/2016 12:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BARTILUCCI,D
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/09/2017
VIOLATION CODE: DUI
ARREST TIME: 22:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:31

SUBJECT'S LAST NAME: CRABTREE

DATE/TIME OF TEST: 03/17/2016 23:54:16

Diagnostics Check	OK	23:51	
Air Blank	0.000	23:51	
Control Test	0.081	23:52	
Air Blank	0.000	23:52	
Subject Sample	# 1	23:52	Subject Test Refused
Breath Volume	0.000	23:52	
Air Blank	0.000	23:53	
Control Test	0.081	23:53	
Air Blank	0.000	23:54	
Diagnostics Check	OK	23:54	

Data Download Date/Time: 03/31/2016 12:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FISSETTE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/09/2017
VIOLATION CODE: DUI
ARREST TIME: 00:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:01

SUBJECT'S LAST NAME: TALAVERA

DATE/TIME OF TEST: 03/18/2016 01:27:50

Diagnostics Check	OK	01:21
Air Blank	0.000	01:22
Control Test	0.081	01:22
Air Blank	0.000	01:22
Subject Sample # 1	0.052	01:23
Breath Volume	1.363	01:23
Air Blank	0.000	01:23
Air Blank	0.000	01:25
Subject Sample # 2	0.050	01:26
Breath Volume	1.722	01:26
Air Blank	0.000	01:26
Control Test	0.081	01:27
Air Blank	0.000	01:27
Diagnostics Check	OK	01:27

Data Download Date/Time: 03/31/2016 12:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ROOT,TROOPER
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 03/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/09/2017
VIOLATION CODE: DUI
ARREST TIME: 04:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:52

SUBJECT'S LAST NAME: RHODEN

DATE/TIME OF TEST: 03/18/2016 06:19:34

Diagnostics Check		OK	06:12
Air Blank		0.000	06:12
Control Test		0.081	06:13
Air Blank		0.000	06:13
Subject Sample	# 1	0.208	06:14
Breath Volume		1.292	06:14
Air Blank		0.000	06:15
Air Blank		0.000	06:17
Subject Sample	# 2	0.218	06:18
Breath Volume		2.175	06:18
Air Blank		0.000	06:18
Control Test		0.081	06:18
Air Blank		0.000	06:19
Diagnostics Check		OK	06:19

Data Download Date/Time: 03/31/2016 12:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCOTT,J,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BARILUCCI,D
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/09/2017
VIOLATION CODE: DUI
ARREST TIME: 23:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:42

SUBJECT'S LAST NAME: COOK

DATE/TIME OF TEST: 03/19/2016 01:06:51

Diagnostics Check	OK	01:03	
Air Blank	0.000	01:04	
Control Test	0.081	01:04	
Air Blank	0.000	01:05	
Subject Sample	# 1	01:05	Subject Test Refused
Breath Volume	0.000	01:05	
Air Blank	0.000	01:05	
Control Test	0.081	01:06	
Air Blank	0.000	01:06	
Diagnostics Check	OK	01:06	

Data Download Date/Time: 03/31/2016 12:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: EANES,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/09/2017
VIOLATION CODE: DUI
ARREST TIME: 22:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:28

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 03/19/2016 23:56:31

Diagnostics Check	OK	23:53	
Air Blank	0.000	23:54	
Control Test	0.080	23:54	
Air Blank	0.000	23:54	
Subject Sample	# 1	23:55	Subject Test Refused
Breath Volume	0.000	23:55	
Air Blank	0.000	23:55	
Control Test	0.081	23:55	
Air Blank	0.000	23:56	
Diagnostics Check	OK	23:56	

Data Download Date/Time: 03/31/2016 12:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MONTGOMERY,K
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 03/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/09/2017
VIOLATION CODE: DUI
ARREST TIME: 22:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:14

SUBJECT'S LAST NAME: ASIALA

DATE/TIME OF TEST: 03/20/2016 00:40:09

Diagnostics Check	OK	00:34
Air Blank	0.000	00:34
Control Test	0.081	00:34
Air Blank	0.000	00:35
Subject Sample # 1	0.000	00:35
Breath Volume	2.792	00:35
Air Blank	0.000	00:36
Air Blank	0.000	00:38
Subject Sample # 2	0.000	00:38
Breath Volume	2.433	00:38
Air Blank	0.000	00:39
Control Test	0.081	00:39
Air Blank	0.000	00:40
Diagnostics Check	OK	00:40

Data Download Date/Time: 03/31/2016 12:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HOWELL,H
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 03/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/09/2017
VIOLATION CODE: DUI
ARREST TIME: 22:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:16

SUBJECT'S LAST NAME: FAILLE

DATE/TIME OF TEST: 03/20/2016 01:42:51

Diagnostics Check	OK	01:36
Air Blank	0.000	01:37
Control Test	0.081	01:37
Air Blank	0.000	01:38
Subject Sample # 1	0.153	01:38
Breath Volume	2.367	01:38
Air Blank	0.000	01:39
Air Blank	0.000	01:40
Subject Sample # 2	0.154	01:41
Breath Volume	2.074	01:41
Air Blank	0.000	01:41
Control Test	0.081	01:42
Air Blank	0.000	01:42
Diagnostics Check	OK	01:42

Data Download Date/Time: 03/31/2016 12:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: EANES,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/09/2017
VIOLATION CODE: DUI
ARREST TIME: 03:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:14

SUBJECT'S LAST NAME: KING

DATE/TIME OF TEST: 03/20/2016 04:40:31

Diagnostics Check	OK	04:34
Air Blank	0.000	04:34
Control Test	0.080	04:34
Air Blank	0.000	04:35
Subject Sample # 1	0.169	04:35
Breath Volume	1.261	04:35
Air Blank	0.000	04:36
Air Blank	0.000	04:38
Subject Sample # 2	0.167	04:39
Breath Volume	1.402	04:39
Air Blank	0.000	04:39
Control Test	0.080	04:39
Air Blank	0.000	04:40
Diagnostics Check	OK	04:40

Data Download Date/Time: 03/31/2016 12:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MONTGOMERY,K
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 03/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/09/2017
VIOLATION CODE: DUI
ARREST TIME: 22:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:51

SUBJECT'S LAST NAME: ZIMMERMAN

DATE/TIME OF TEST: 03/21/2016 00:17:46

Diagnostics Check	OK	00:11
Air Blank	0.000	00:12
Control Test	0.080	00:12
Air Blank	0.000	00:13
Subject Sample # 1	0.000	00:13
Breath Volume	2.507	00:13
Air Blank	0.000	00:14
Air Blank	0.000	00:16
Subject Sample # 2	0.000	00:16
Breath Volume	2.726	00:16
Air Blank	0.000	00:16
Control Test	0.081	00:17
Air Blank	0.000	00:17
Diagnostics Check	OK	00:17

Data Download Date/Time: 03/31/2016 12:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BYRNE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/09/2017
VIOLATION CODE: DUI
ARREST TIME: 01:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: MCCULLOCH

DATE/TIME OF TEST: 03/21/2016 03:07:26

Diagnostics Check		OK	02:55	
Air Blank		0.000	02:56	
Control Test		0.081	02:56	
Air Blank		0.000	02:57	
Subject Sample	# 1	0.199	03:00	Volume Not Met
Breath Volume		0.156	03:00	
Air Blank		0.000	03:00	
Air Blank		0.000	03:02	
Subject Sample	# 2	0.166	03:05	Volume Not Met
Breath Volume		1.050	03:05	
Air Blank		0.000	03:06	
Control Test		0.080	03:06	
Air Blank		0.000	03:07	
Diagnostics Check		OK	03:07	

Data Download Date/Time: 03/31/2016 12:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BYRNE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/09/2017
VIOLATION CODE: DUI
ARREST TIME: 01:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: MCCULLOCH

DATE/TIME OF TEST: 03/21/2016 03:26:21

Diagnostics Check		OK	03:09	
Air Blank		0.000	03:09	
Control Test		0.081	03:09	
Air Blank		0.000	03:10	
Subject Sample	# 1	0.184	03:13	Slope Not Level
Breath Volume		1.515	03:13	
Air Blank		0.000	03:14	
Air Blank		0.000	03:15	
Subject Sample	# 2	0.208	03:19	
Breath Volume		1.265	03:19	
Air Blank		0.000	03:19	
Air Blank		0.000	03:21	
Subject Sample	# 3	0.181	03:24	Volume Not Met
Breath Volume		0.000	03:24	
Air Blank		0.000	03:25	
Control Test		0.080	03:25	
Air Blank		0.000	03:26	
Diagnostics Check		OK	03:26	

Data Download Date/Time: 03/31/2016 12:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: EANES,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/09/2017
VIOLATION CODE: DUI
ARREST TIME: 22:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: SWEETING

DATE/TIME OF TEST: 03/23/2016 00:18:50

Diagnostics Check	OK	00:16	
Air Blank	0.000	00:16	
Control Test	0.082	00:16	
Air Blank	0.000	00:17	
Subject Sample	# 1	00:17	Subject Test Refused
Breath Volume	0.000	00:17	
Air Blank	0.000	00:17	
Control Test	0.081	00:18	
Air Blank	0.000	00:18	
Diagnostics Check	OK	00:18	

Data Download Date/Time: 03/31/2016 12:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: KERSHAW,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/09/2017
VIOLATION CODE: DUI
ARREST TIME: 04:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:58

SUBJECT'S LAST NAME: CALDER

DATE/TIME OF TEST: 03/23/2016 05:21:38

Diagnostics Check	OK	05:18	
Air Blank	0.000	05:19	
Control Test	0.083	05:19	
Air Blank	0.000	05:20	
Subject Sample	# 1	05:20	Subject Test Refused
Breath Volume	0.000	05:20	
Air Blank	0.000	05:20	
Control Test	0.081	05:21	
Air Blank	0.000	05:21	
Diagnostics Check	OK	05:21	

Data Download Date/Time: 03/31/2016 12:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BYRNE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/09/2017
VIOLATION CODE: DUI
ARREST TIME: 22:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:49

SUBJECT'S LAST NAME: PIACENTINO

DATE/TIME OF TEST: 03/24/2016 00:16:06

Diagnostics Check	OK	00:10
Air Blank	0.000	00:10
Control Test	0.081	00:10
Air Blank	0.000	00:11
Subject Sample # 1	0.146	00:11
Breath Volume	2.781	00:11
Air Blank	0.000	00:12
Air Blank	0.000	00:14
Subject Sample # 2	0.148	00:14
Breath Volume	2.578	00:14
Air Blank	0.000	00:15
Control Test	0.080	00:15
Air Blank	0.000	00:16
Diagnostics Check	OK	00:16

Data Download Date/Time: 03/31/2016 12:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MONTGOMERY,K
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/09/2017
VIOLATION CODE: DUI
ARREST TIME: 23:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:43

SUBJECT'S LAST NAME: SATCHER

DATE/TIME OF TEST: 03/24/2016 01:09:52

Diagnostics Check	OK	01:03
Air Blank	0.000	01:03
Control Test	0.081	01:04
Air Blank	0.000	01:04
Subject Sample # 1	0.221	01:05
Breath Volume	3.578	01:05
Air Blank	0.000	01:05
Air Blank	0.000	01:07
Subject Sample # 2	0.212	01:08
Breath Volume	3.847	01:08
Air Blank	0.000	01:08
Control Test	0.081	01:09
Air Blank	0.000	01:09
Diagnostics Check	OK	01:09

Data Download Date/Time: 03/31/2016 12:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: PILTZ,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/09/2017
VIOLATION CODE: DUI
ARREST TIME: 03:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:22

SUBJECT'S LAST NAME: BLUE

DATE/TIME OF TEST: 03/24/2016 05:45:57

Diagnostics Check	OK	05:43	
Air Blank	0.000	05:43	
Control Test	0.081	05:43	
Air Blank	0.000	05:44	
Subject Sample	# 1	05:44	Subject Test Refused
Breath Volume	0.000	05:44	
Air Blank	0.000	05:45	
Control Test	0.080	05:45	
Air Blank	0.000	05:45	
Diagnostics Check	OK	05:45	

Data Download Date/Time: 03/31/2016 12:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: STANLEY,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/09/2017
VIOLATION CODE: DUI
ARREST TIME: 22:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:16

SUBJECT'S LAST NAME: WRIGHT

DATE/TIME OF TEST: 03/25/2016 23:40:07

Diagnostics Check	OK	23:37	
Air Blank	0.000	23:37	
Control Test	0.079	23:37	
Air Blank	0.000	23:38	
Subject Sample	# 1	23:38	Subject Test Refused
Breath Volume	0.000	23:38	
Air Blank	0.000	23:39	
Control Test	0.081	23:39	
Air Blank	0.000	23:40	
Diagnostics Check	OK	23:40	

Data Download Date/Time: 03/31/2016 12:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FISSETT,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/09/2017
VIOLATION CODE: DUI
ARREST TIME: 23:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:21

SUBJECT'S LAST NAME: SPOFFORD

DATE/TIME OF TEST: 03/26/2016 00:47:17

Diagnostics Check	OK	00:41
Air Blank	0.000	00:41
Control Test	0.081	00:42
Air Blank	0.000	00:42
Subject Sample # 1	0.212	00:42
Breath Volume	2.488	00:42
Air Blank	0.000	00:43
Air Blank	0.000	00:45
Subject Sample # 2	0.205	00:45
Breath Volume	2.410	00:45
Air Blank	0.000	00:46
Control Test	0.080	00:46
Air Blank	0.000	00:47
Diagnostics Check	OK	00:47

Data Download Date/Time: 03/31/2016 12:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: KLING,D
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 03/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/09/2017
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: FUSCO

DATE/TIME OF TEST: 03/26/2016 03:37:46

Diagnostics Check		OK	03:30
Air Blank		0.000	03:31
Control Test		0.081	03:31
Air Blank		0.000	03:32
Subject Sample	# 1	0.138	03:32
Breath Volume		1.308	03:32
Air Blank		0.000	03:33
Air Blank		0.000	03:35
Subject Sample	# 2	0.149	03:36
Breath Volume		1.460	03:36
Air Blank		0.000	03:36
Control Test		0.080	03:37
Air Blank		0.000	03:37
Diagnostics Check		OK	03:37

Data Download Date/Time: 03/31/2016 12:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DAVIS,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/09/2017
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:32

SUBJECT'S LAST NAME: TRAN

DATE/TIME OF TEST: 03/26/2016 04:58:17

Diagnostics Check	OK	04:52
Air Blank	0.000	04:52
Control Test	0.080	04:53
Air Blank	0.000	04:53
Subject Sample # 1	0.212	04:53
Breath Volume	2.953	04:53
Air Blank	0.000	04:54
Air Blank	0.000	04:56
Subject Sample # 2	0.204	04:56
Breath Volume	2.089	04:56
Air Blank	0.000	04:57
Control Test	0.080	04:57
Air Blank	0.000	04:58
Diagnostics Check	OK	04:58

Data Download Date/Time: 03/31/2016 12:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: GARCIA,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/09/2017
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: MCDONALD

DATE/TIME OF TEST: 03/27/2016 01:59:41

Diagnostics Check	OK	01:53
Air Blank	0.000	01:53
Control Test	0.081	01:53
Air Blank	0.000	01:54
Subject Sample # 1	0.092	01:55
Breath Volume	1.214	01:55
Air Blank	0.000	01:55
Air Blank	0.000	01:57
Subject Sample # 2	0.096	01:58
Breath Volume	1.640	01:58
Air Blank	0.000	01:58
Control Test	0.081	01:59
Air Blank	0.000	01:59
Diagnostics Check	OK	01:59

Data Download Date/Time: 03/31/2016 12:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: TERRELL,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/09/2017
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:36

SUBJECT'S LAST NAME: PAW

DATE/TIME OF TEST: 03/27/2016 03:07:23

Diagnostics Check	OK	03:01
Air Blank	0.000	03:01
Control Test	0.081	03:01
Air Blank	0.000	03:02
Subject Sample # 1	0.187	03:03
Breath Volume	2.433	03:03
Air Blank	0.000	03:03
Air Blank	0.000	03:05
Subject Sample # 2	0.184	03:05
Breath Volume	1.906	03:05
Air Blank	0.000	03:06
Control Test	0.080	03:06
Air Blank	0.000	03:07
Diagnostics Check	OK	03:07

Data Download Date/Time: 03/31/2016 12:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: PETERS,R
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 03/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/09/2017
VIOLATION CODE: DUI
ARREST TIME: 01:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:15

SUBJECT'S LAST NAME: COLEMAN

DATE/TIME OF TEST: 03/27/2016 04:48:30

Diagnostics Check	OK	04:45	
Air Blank	0.000	04:46	
Control Test	0.081	04:46	
Air Blank	0.000	04:46	
Subject Sample	# 1	04:47	Subject Test Refused
Breath Volume	0.000	04:47	
Air Blank	0.000	04:47	
Control Test	0.080	04:47	
Air Blank	0.000	04:48	
Diagnostics Check	OK	04:48	

Data Download Date/Time: 03/31/2016 12:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: PETERS,R
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 03/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/09/2017
VIOLATION CODE: DUI
ARREST TIME: 01:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:11

SUBJECT'S LAST NAME: COLEMAN

DATE/TIME OF TEST: 03/27/2016 05:37:54

Diagnostics Check	OK	05:31
Air Blank	0.000	05:32
Control Test	0.080	05:32
Air Blank	0.000	05:32
Subject Sample # 1	0.075	05:33
Breath Volume	1.675	05:33
Air Blank	0.000	05:33
Air Blank	0.000	05:35
Subject Sample # 2	0.077	05:36
Breath Volume	2.011	05:36
Air Blank	0.000	05:36
Control Test	0.080	05:37
Air Blank	0.000	05:37
Diagnostics Check	OK	05:37

Data Download Date/Time: 03/31/2016 12:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: COPPOLA,S
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 03/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/09/2017
VIOLATION CODE: DUI
ARREST TIME: 23:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:32

SUBJECT'S LAST NAME: MCCLENDON

DATE/TIME OF TEST: 03/29/2016 02:59:31

Diagnostics Check	OK	02:52
Air Blank	0.000	02:53
Control Test	0.080	02:53
Air Blank	0.000	02:54
Subject Sample # 1	0.043	02:54
Breath Volume	1.312	02:54
Air Blank	0.000	02:55
Air Blank	0.000	02:57
Subject Sample # 2	0.043	02:58
Breath Volume	1.824	02:58
Air Blank	0.000	02:58
Control Test	0.080	02:58
Air Blank	0.000	02:59
Diagnostics Check	OK	02:59

Data Download Date/Time: 03/31/2016 12:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BYRNE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/31/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/09/2017
VIOLATION CODE: DUI
ARREST TIME: 21:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:24

SUBJECT'S LAST NAME: CONNER

DATE/TIME OF TEST: 03/31/2016 23:50:14

Diagnostics Check	OK	23:44
Air Blank	0.000	23:44
Control Test	0.080	23:44
Air Blank	0.000	23:45
Subject Sample # 1	0.266	23:45
Breath Volume	3.214	23:45
Air Blank	0.000	23:46
Air Blank	0.000	23:48
Subject Sample # 2	0.261	23:48
Breath Volume	3.191	23:48
Air Blank	0.000	23:49
Control Test	0.081	23:49
Air Blank	0.000	23:50
Diagnostics Check	OK	23:50

Data Download Date/Time: 04/15/2016 12:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALLIDAY,ROBERT,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MARRERO,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:33

SUBJECT'S LAST NAME: BROOKS

DATE/TIME OF TEST: 03/03/2016 01:55:58

Diagnostics Check	OK	01:53	
Air Blank	0.000	01:53	
Control Test	0.080	01:53	
Air Blank	0.000	01:54	
Subject Sample	# 1	01:54	Subject Test Refused
Breath Volume	0.000	01:54	
Air Blank	0.000	01:55	
Control Test	0.080	01:55	
Air Blank	0.000	01:55	
Diagnostics Check	OK	01:55	

Data Download Date/Time: 03/15/2016 13:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALLIDAY,ROBERT,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BRYANT,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:37

SUBJECT'S LAST NAME: BELLAMY

DATE/TIME OF TEST: 03/03/2016 03:04:32

Diagnostics Check	OK	02:57
Air Blank	0.000	02:58
Control Test	0.079	02:58
Air Blank	0.000	02:59
Subject Sample # 1	0.061	03:00
Breath Volume	1.667	03:00
Air Blank	0.000	03:00
Air Blank	0.000	03:02
Subject Sample # 2	0.063	03:03
Breath Volume	2.207	03:03
Air Blank	0.000	03:03
Control Test	0.079	03:03
Air Blank	0.000	03:04
Diagnostics Check	OK	03:04

Data Download Date/Time: 03/15/2016 13:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THIGPEN,J,T
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BROWN,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 19:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:34

SUBJECT'S LAST NAME: CIFUENTES-ZAMBRANO

DATE/TIME OF TEST: 03/04/2016 22:10:58

Diagnostics Check	OK	22:04
Air Blank	0.000	22:04
Control Test	0.080	22:05
Air Blank	0.000	22:05
Subject Sample # 1	0.146	22:06
Breath Volume	1.675	22:06
Air Blank	0.000	22:07
Air Blank	0.000	22:08
Subject Sample # 2	0.156	22:09
Breath Volume	2.390	22:09
Air Blank	0.000	22:10
Control Test	0.079	22:10
Air Blank	0.000	22:10
Diagnostics Check	OK	22:10

Data Download Date/Time: 03/15/2016 13:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY,SHANON,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: TERRELL,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: MAYO

DATE/TIME OF TEST: 03/09/2016 01:21:51

Diagnostics Check	OK	01:17	
Air Blank	0.000	01:17	
Control Test	0.079	01:18	
Air Blank	0.000	01:18	
Subject Sample	# 1	01:20	Subject Test Refused
Breath Volume	0.000	01:20	
Air Blank	0.000	01:20	
Control Test	0.080	01:21	
Air Blank	0.000	01:21	
Diagnostics Check	OK	01:21	

Data Download Date/Time: 03/15/2016 13:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY,SHANON,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: KEEN,C.E.
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 02/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:02

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 03/09/2016 04:26:24

Diagnostics Check	OK	04:23	
Air Blank	0.000	04:24	
Control Test	0.080	04:24	
Air Blank	0.000	04:24	
Subject Sample	# 1	04:25	Subject Test Refused
Breath Volume	0.000	04:25	
Air Blank	0.000	04:25	
Control Test	0.080	04:25	
Air Blank	0.000	04:26	
Diagnostics Check	OK	04:26	

Data Download Date/Time: 03/15/2016 13:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THIGPEN,J,T
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BYRNE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 19:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:47

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 03/12/2016 22:17:17

Diagnostics Check	OK	22:10
Air Blank	0.000	22:11
Control Test	0.078	22:11
Air Blank	0.000	22:12
Subject Sample # 1	0.174	22:12
Breath Volume	3.574	22:12
Air Blank	0.000	22:13
Air Blank	0.000	22:15
Subject Sample # 2	0.174	22:15
Breath Volume	3.816	22:15
Air Blank	0.000	22:16
Control Test	0.078	22:16
Air Blank	0.000	22:17
Diagnostics Check	OK	22:17

Data Download Date/Time: 03/15/2016 13:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THIGPEN,J,T
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: GARCIA,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A2
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 18:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:52

SUBJECT'S LAST NAME: FULFORD

DATE/TIME OF TEST: 03/17/2016 20:27:50

Diagnostics Check	OK	20:20
Air Blank	0.000	20:20
Control Test	0.079	20:20
Air Blank	0.000	20:21
Subject Sample # 1	0.238	20:22
Breath Volume	2.320	20:22
Air Blank	0.000	20:23
Air Blank	0.000	20:25
Subject Sample # 2	0.253	20:26
Breath Volume	2.015	20:26
Air Blank	0.000	20:26
Control Test	0.079	20:27
Air Blank	0.000	20:27
Diagnostics Check	OK	20:27

Data Download Date/Time: 03/31/2016 12:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THIGPEN,J,T
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: TERRELL,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A2
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:55

SUBJECT'S LAST NAME: ROBINETTE

DATE/TIME OF TEST: 03/17/2016 22:23:18

Diagnostics Check	OK	22:16
Air Blank	0.000	22:17
Control Test	0.080	22:17
Air Blank	0.000	22:18
Subject Sample # 1	0.189	22:18
Breath Volume	2.582	22:18
Air Blank	0.000	22:19
Air Blank	0.000	22:21
Subject Sample # 2	0.196	22:21
Breath Volume	2.503	22:21
Air Blank	0.000	22:22
Control Test	0.080	22:22
Air Blank	0.000	22:23
Diagnostics Check	OK	22:23

Data Download Date/Time: 03/31/2016 12:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY,SHANON,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: TERRELL,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A2
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 03/18/2016 00:27:12

Diagnostics Check	OK	00:26	
Air Blank		00:26	RFI Detect
Air Blank	0.000	00:27	

Data Download Date/Time: 03/31/2016 12:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY,SHANON,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: TERRRELL,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A2
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 03/18/2016 00:34:34

Diagnostics Check	OK	00:28
Air Blank	0.000	00:28
Control Test	0.080	00:28
Air Blank	0.000	00:29
Subject Sample # 1	0.000	00:30
Breath Volume	2.097	00:30
Air Blank	0.000	00:30
Air Blank	0.000	00:32
Subject Sample # 2	0.000	00:33
Breath Volume	2.281	00:33
Air Blank	0.000	00:33
Control Test	0.081	00:33
Air Blank	0.000	00:34
Diagnostics Check	OK	00:34

Data Download Date/Time: 03/31/2016 12:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY, SHANON, E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: TERRELL, T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A2
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 03/18/2016 03:31:50

Diagnostics Check	OK	03:31	
Air Blank		03:31	RFI Detect
Air Blank	0.000	03:31	

Data Download Date/Time: 03/31/2016 12:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY,SHANO,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: TERRELL,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A2
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 03/18/2016 03:39:13

Diagnostics Check	OK	03:32
Air Blank	0.000	03:33
Control Test	0.080	03:33
Air Blank	0.000	03:34
Subject Sample # 1	0.076	03:34
Breath Volume	1.707	03:34
Air Blank	0.000	03:35
Air Blank	0.000	03:37
Subject Sample # 2	0.075	03:37
Breath Volume	2.320	03:37
Air Blank	0.000	03:38
Control Test	0.079	03:38
Air Blank	0.000	03:39
Diagnostics Check	OK	03:39

Data Download Date/Time: 03/31/2016 12:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCOTT,J,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: GAZIC,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A2
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: MARTISEK

DATE/TIME OF TEST: 03/19/2016 02:53:56

Diagnostics Check	OK	02:51	
Air Blank	0.000	02:51	
Control Test	0.080	02:51	
Air Blank	0.000	02:52	
Subject Sample	# 1	02:52	Subject Test Refused
Breath Volume	0.000	02:52	
Air Blank	0.000	02:53	
Control Test	0.080	02:53	
Air Blank	0.000	02:53	
Diagnostics Check	OK	02:53	

Data Download Date/Time: 03/31/2016 12:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCOTT,J,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DANZIE,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A2
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 03/19/2016 04:03:12

Diagnostics Check	OK	04:00	
Air Blank	0.000	04:00	
Control Test	0.080	04:01	
Air Blank	0.000	04:01	
Subject Sample	# 1	04:01	Subject Test Refused
Breath Volume	0.000	04:01	
Air Blank	0.000	04:02	
Control Test	0.080	04:02	
Air Blank	0.000	04:03	
Diagnostics Check	OK	04:03	

Data Download Date/Time: 03/31/2016 12:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCOTT,J,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: THOMPSON,M
ARRESTING OFFICER AGENCY: Atlantic Beach PD
LAST AGENCY INSPECTION: 03/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A2
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:56

SUBJECT'S LAST NAME: JEAN-BART

DATE/TIME OF TEST: 03/19/2016 05:26:38

Diagnostics Check	OK	05:23	
Air Blank	0.000	05:24	
Control Test	0.080	05:24	
Air Blank	0.000	05:25	
Subject Sample	# 1	05:25	Subject Test Refused
Breath Volume	0.000	05:25	
Air Blank	0.000	05:25	
Control Test	0.079	05:26	
Air Blank	0.000	05:26	
Diagnostics Check	OK	05:26	

Data Download Date/Time: 03/31/2016 12:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THIGPEN,J,T
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: TERRELL,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A2
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 17:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:15

SUBJECT'S LAST NAME: PIERCE

DATE/TIME OF TEST: 03/26/2016 18:40:00

Diagnostics Check	OK	18:36	
Air Blank	0.000	18:37	
Control Test	0.080	18:37	
Air Blank	0.000	18:37	
Subject Sample	# 1	18:38	Subject Test Refused
Breath Volume	0.000	18:38	
Air Blank	0.000	18:39	
Control Test	0.080	18:39	
Air Blank	0.000	18:39	
Diagnostics Check	OK	18:40	

Data Download Date/Time: 03/31/2016 12:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THIGPEN,J,T
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: TERRELL,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A2
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:21

SUBJECT'S LAST NAME: HOLLINGSWORTH

DATE/TIME OF TEST: 03/26/2016 22:48:20

Diagnostics Check	OK	22:41
Air Blank	0.000	22:42
Control Test	0.078	22:42
Air Blank	0.000	22:43
Subject Sample # 1	0.151	22:43
Breath Volume	3.285	22:43
Air Blank	0.000	22:44
Air Blank	0.000	22:46
Subject Sample # 2	0.153	22:46
Breath Volume	2.917	22:46
Air Blank	0.000	22:47
Control Test	0.079	22:47
Air Blank	0.000	22:48
Diagnostics Check	OK	22:48

Data Download Date/Time: 03/31/2016 12:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THIGPEN,J,T
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: EANES,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A2
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 18:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:35

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 03/29/2016 20:01:46

Diagnostics Check	OK	19:55
Air Blank	0.000	19:55
Control Test	0.079	19:56
Air Blank	0.000	19:56
Subject Sample # 1	0.000	19:57
Breath Volume	2.027	19:57
Air Blank	0.000	19:57
Air Blank	0.000	19:59
Subject Sample # 2	0.000	20:00
Breath Volume	2.992	20:00
Air Blank	0.000	20:00
Control Test	0.080	20:01
Air Blank	0.000	20:01
Diagnostics Check	OK	20:01

Data Download Date/Time: 03/31/2016 12:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAKER, TYLER, R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BROWN, W
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 16:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:05

SUBJECT'S LAST NAME: REED

DATE/TIME OF TEST: 03/01/2016 17:31:57

Diagnostics Check	OK	17:25
Air Blank	0.000	17:25
Control Test	0.079	17:26
Air Blank	0.000	17:26
Subject Sample # 1	0.126	17:27
Breath Volume	2.867	17:27
Air Blank	0.000	17:27
Air Blank	0.000	17:29
Subject Sample # 2	0.128	17:30
Breath Volume	3.109	17:30
Air Blank	0.000	17:31
Control Test	0.079	17:31
Air Blank	0.000	17:31
Diagnostics Check	OK	17:31

Data Download Date/Time: 03/15/2016 12:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRUETT,RICHARD,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: EANES,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: BIVINS

DATE/TIME OF TEST: 03/02/2016 01:36:48

Diagnostics Check	OK	01:30
Air Blank	0.000	01:30
Control Test	0.079	01:30
Air Blank	0.000	01:31
Subject Sample # 1	0.112	01:31
Breath Volume	1.523	01:31
Air Blank	0.000	01:32
Air Blank	0.000	01:34
Subject Sample # 2	0.115	01:35
Breath Volume	2.761	01:35
Air Blank	0.000	01:35
Control Test	0.079	01:36
Air Blank	0.000	01:36
Diagnostics Check	OK	01:36

Data Download Date/Time: 03/15/2016 12:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRUETT,RICHARD,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: KERSHAW,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:25

SUBJECT'S LAST NAME: AKINS

DATE/TIME OF TEST: 03/02/2016 04:52:52

Diagnostics Check	OK	04:46
Air Blank	0.000	04:46
Control Test	0.079	04:47
Air Blank	0.000	04:47
Subject Sample # 1	0.098	04:48
Breath Volume	2.101	04:48
Air Blank	0.000	04:48
Air Blank	0.000	04:50
Subject Sample # 2	0.097	04:51
Breath Volume	2.460	04:51
Air Blank	0.000	04:51
Control Test	0.079	04:52
Air Blank	0.000	04:52
Diagnostics Check	OK	04:52

Data Download Date/Time: 03/15/2016 12:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRUETT,RICHARD,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: EANES,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:34

SUBJECT'S LAST NAME: MCGHEE

DATE/TIME OF TEST: 03/03/2016 00:01:38

Diagnostics Check	OK	23:54
Air Blank	0.000	23:54
Control Test	0.080	23:55
Air Blank	0.000	23:55
Subject Sample # 2	0.147	23:57
Breath Volume	3.054	23:57
Air Blank	0.000	23:57
Air Blank	0.000	23:59
Subject Sample # 1	0.143	00:00
Breath Volume	1.667	00:00
Air Blank	0.000	00:00
Control Test	0.078	00:01
Air Blank	0.000	00:01
Diagnostics Check	OK	00:01

Data Download Date/Time: 03/15/2016 12:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRUETT,RICHARD,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BAIS
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: ELLIS

DATE/TIME OF TEST: 03/03/2016 02:14:43

Diagnostics Check	OK	02:08
Air Blank	0.000	02:08
Control Test	0.079	02:08
Air Blank	0.000	02:09
Subject Sample # 1	0.187	02:10
Breath Volume	1.121	02:10
Air Blank	0.000	02:10
Air Blank	0.000	02:12
Subject Sample # 2	0.182	02:13
Breath Volume	2.003	02:13
Air Blank	0.000	02:13
Control Test	0.079	02:14
Air Blank	0.000	02:14
Diagnostics Check	OK	02:14

Data Download Date/Time: 03/15/2016 12:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTIN,BRADEN,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MYNIER,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:34

SUBJECT'S LAST NAME: WEINBURG

DATE/TIME OF TEST: 03/04/2016 06:00:59

Diagnostics Check	OK	05:54
Air Blank	0.000	05:54
Control Test	0.080	05:55
Air Blank	0.000	05:55
Subject Sample # 1	0.129	05:56
Breath Volume	1.375	05:56
Air Blank	0.000	05:57
Air Blank	0.000	05:59
Subject Sample # 2	0.125	05:59
Breath Volume	1.214	05:59
Air Blank	0.000	06:00
Control Test	0.079	06:00
Air Blank	0.000	06:00
Diagnostics Check	OK	06:00

Data Download Date/Time: 03/15/2016 12:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALLIDAY,ROBERT,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BARTILUCCI,D
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:32

SUBJECT'S LAST NAME: ENGLE

DATE/TIME OF TEST: 03/05/2016 22:58:30

Diagnostics Check	OK	22:52
Air Blank	0.000	22:52
Control Test	0.079	22:52
Air Blank	0.000	22:53
Subject Sample # 1	0.244	22:54
Breath Volume	2.718	22:54
Air Blank	0.000	22:54
Air Blank	0.000	22:56
Subject Sample # 2	0.241	22:56
Breath Volume	2.476	22:56
Air Blank	0.000	22:57
Control Test	0.079	22:57
Air Blank	0.000	22:58
Diagnostics Check	OK	22:58

Data Download Date/Time: 03/15/2016 12:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALLIDAY,ROBERT,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: GARCIA,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:37

SUBJECT'S LAST NAME: SELDEN

DATE/TIME OF TEST: 03/06/2016 04:04:17

Diagnostics Check	OK	03:58
Air Blank	0.000	03:58
Control Test	0.079	03:58
Air Blank	0.000	03:59
Subject Sample # 1	0.265	03:59
Breath Volume	3.339	03:59
Air Blank	0.000	04:00
Air Blank	0.000	04:02
Subject Sample # 2	0.254	04:02
Breath Volume	2.363	04:02
Air Blank	0.000	04:03
Control Test	0.079	04:03
Air Blank	0.000	04:04
Diagnostics Check	OK	04:04

Data Download Date/Time: 03/15/2016 12:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALLIDAY,ROBERT,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HOOVER,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:22

SUBJECT'S LAST NAME: CHANCE

DATE/TIME OF TEST: 03/06/2016 04:46:07

Diagnostics Check	OK	04:43	
Air Blank	0.000	04:43	
Control Test	0.080	04:43	
Air Blank	0.000	04:44	
Subject Sample	# 1	04:44	Subject Test Refused
Breath Volume	0.000	04:44	
Air Blank	0.000	04:45	
Control Test	0.080	04:45	
Air Blank	0.000	04:46	
Diagnostics Check	OK	04:46	

Data Download Date/Time: 03/15/2016 12:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRUETT,RICHARD,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WYATT,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 05:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:43

SUBJECT'S LAST NAME: GALICIA

DATE/TIME OF TEST: 03/06/2016 07:13:11

Diagnostics Check	OK	07:05
Air Blank	0.000	07:06
Control Test	0.080	07:06
Air Blank	0.000	07:07
Subject Sample # 1	0.145	07:08
Breath Volume	2.128	07:08
Air Blank	0.000	07:08
Air Blank	0.000	07:10
Subject Sample # 2	0.144	07:11
Breath Volume	2.921	07:11
Air Blank	0.000	07:12
Control Test	0.079	07:12
Air Blank	0.000	07:13
Diagnostics Check	OK	07:13

Data Download Date/Time: 03/15/2016 12:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALLIDAY,ROBERT,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHITE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:52

SUBJECT'S LAST NAME: CARDONA

DATE/TIME OF TEST: 03/07/2016 00:18:37

Diagnostics Check	OK	00:12
Air Blank	0.000	00:13
Control Test	0.080	00:13
Air Blank	0.000	00:14
Subject Sample # 1	0.010	00:14
Breath Volume	1.820	00:14
Air Blank	0.000	00:14
Air Blank	0.000	00:16
Subject Sample # 2	0.009	00:17
Breath Volume	2.031	00:17
Air Blank	0.000	00:17
Control Test	0.080	00:18
Air Blank	0.000	00:18
Diagnostics Check	OK	00:18

Data Download Date/Time: 03/15/2016 12:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALLIDAY,ROBERT,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: KLING,D
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 02/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:54

SUBJECT'S LAST NAME: STEWART

DATE/TIME OF TEST: 03/07/2016 02:20:56

Diagnostics Check	OK	02:14
Air Blank	0.000	02:15
Control Test	0.080	02:15
Air Blank	0.000	02:16
Subject Sample # 1	0.147	02:16
Breath Volume	3.546	02:16
Air Blank	0.000	02:17
Air Blank	0.000	02:19
Subject Sample # 2	0.141	02:19
Breath Volume	2.531	02:19
Air Blank	0.000	02:20
Control Test	0.080	02:20
Air Blank	0.000	02:20
Diagnostics Check	OK	02:20

Data Download Date/Time: 03/15/2016 12:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALLIDAY,ROBERT,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FISETTE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: ALVAREZ

DATE/TIME OF TEST: 03/07/2016 03:38:59

Diagnostics Check	OK	03:36	
Air Blank	0.000	03:36	
Control Test	0.080	03:36	
Air Blank	0.000	03:37	
Subject Sample	# 1	03:37	Subject Test Refused
Breath Volume	0.000	03:37	
Air Blank	0.000	03:38	
Control Test	0.080	03:38	
Air Blank	0.000	03:38	
Diagnostics Check	OK	03:38	

Data Download Date/Time: 03/15/2016 12:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAKER, TYLER, R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: TERRELL, T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 17:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:40

SUBJECT'S LAST NAME: FUSSELL

DATE/TIME OF TEST: 03/09/2016 19:06:02

Diagnostics Check	OK	19:03	
Air Blank	0.000	19:03	
Control Test	0.078	19:04	
Air Blank	0.000	19:04	
Subject Sample	# 1	19:04	Subject Test Refused
Breath Volume	0.000	19:04	
Air Blank	0.000	19:05	
Control Test	0.080	19:05	
Air Blank	0.000	19:05	
Diagnostics Check	OK	19:06	

Data Download Date/Time: 03/15/2016 12:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON, BRADEN, J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: GARCIA, J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 03/18/2016 01:23:14

Diagnostics Check	OK	01:20	
Air Blank	0.000	01:20	
Control Test	0.078	01:21	
Air Blank	0.000	01:21	
Subject Sample	# 1	01:21	Subject Test Refused
Breath Volume	0.000	01:21	
Air Blank	0.000	01:22	
Control Test	0.079	01:22	
Air Blank	0.000	01:23	
Diagnostics Check	OK	01:23	

Data Download Date/Time: 03/31/2016 12:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON, BRADEN, J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FISETTE, J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:42

SUBJECT'S LAST NAME: WELLS

DATE/TIME OF TEST: 03/18/2016 04:06:01

Diagnostics Check	OK	04:03	
Air Blank	0.000	04:03	
Control Test	0.079	04:03	
Air Blank	0.000	04:04	
Subject Sample	# 1	04:04	Subject Test Refused
Breath Volume	0.000	04:04	
Air Blank	0.000	04:05	
Control Test	0.080	04:05	
Air Blank	0.000	04:05	
Diagnostics Check	OK	04:06	

Data Download Date/Time: 03/31/2016 12:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAKER, TYLER, R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: GARCIA, J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: WEINDORF

DATE/TIME OF TEST: 03/18/2016 21:27:10

Diagnostics Check	OK	21:20
Air Blank	0.000	21:21
Control Test	0.079	21:21
Air Blank	0.000	21:22
Subject Sample # 1	0.072	21:22
Breath Volume	2.652	21:22
Air Blank	0.000	21:23
Air Blank	0.000	21:25
Subject Sample # 2	0.069	21:25
Breath Volume	2.699	21:25
Air Blank	0.000	21:26
Control Test	0.079	21:26
Air Blank	0.000	21:27
Diagnostics Check	OK	21:27

Data Download Date/Time: 03/31/2016 12:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON, BRADEN, J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: TERRELL, T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:05

SUBJECT'S LAST NAME: SOUDERS

DATE/TIME OF TEST: 03/18/2016 23:32:13

Diagnostics Check	OK	23:25
Air Blank	0.000	23:26
Control Test	0.079	23:26
Air Blank	0.000	23:26
Subject Sample # 1	0.000	23:27
Breath Volume	1.925	23:27
Air Blank	0.000	23:27
Air Blank	0.000	23:29
Subject Sample # 2	0.000	23:30
Breath Volume	2.261	23:30
Air Blank	0.000	23:31
Control Test	0.079	23:31
Air Blank	0.000	23:32
Diagnostics Check	OK	23:32

Data Download Date/Time: 03/31/2016 12:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON, BRADEN, J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FISETTE, J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:09

SUBJECT'S LAST NAME: OSORIO RAMIREZ

DATE/TIME OF TEST: 03/19/2016 00:36:05

Diagnostics Check	OK	00:29
Air Blank	0.000	00:30
Control Test	0.079	00:30
Air Blank	0.000	00:30
Subject Sample # 1	0.098	00:31
Breath Volume	1.472	00:31
Air Blank	0.000	00:32
Air Blank	0.000	00:34
Subject Sample # 2	0.097	00:34
Breath Volume	1.105	00:34
Air Blank	0.000	00:35
Control Test	0.079	00:35
Air Blank	0.000	00:36
Diagnostics Check	OK	00:36

Data Download Date/Time: 03/31/2016 12:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON, BRADEN, J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: GARCIA, J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: MENDEZ

DATE/TIME OF TEST: 03/19/2016 01:56:51

Diagnostics Check	OK	01:50
Air Blank	0.000	01:50
Control Test	0.079	01:51
Air Blank	0.000	01:51
Subject Sample # 1	0.147	01:52
Breath Volume	2.683	01:52
Air Blank	0.000	01:52
Air Blank	0.000	01:54
Subject Sample # 2	0.149	01:55
Breath Volume	1.757	01:55
Air Blank	0.000	01:55
Control Test	0.079	01:56
Air Blank	0.000	01:56
Diagnostics Check	OK	01:56

Data Download Date/Time: 03/31/2016 12:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON, BRADEN, J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: TERRELL, T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:08

SUBJECT'S LAST NAME: DEVINCENTIS

DATE/TIME OF TEST: 03/19/2016 02:42:04

Diagnostics Check	OK	02:32
Air Blank	0.000	02:32
Control Test	0.079	02:33
Air Blank	0.000	02:33
Subject Sample # 1	0.064	02:35
Breath Volume	1.417	02:35
Air Blank	0.000	02:36
Air Blank	0.000	02:38
Subject Sample # 2	0.062	02:40
Breath Volume	1.648	02:40
Air Blank	0.000	02:41
Control Test	0.078	02:41
Air Blank	0.000	02:41
Diagnostics Check	OK	02:42

Data Download Date/Time: 03/31/2016 12:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON, BRADEN, J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MONTGOMERY, K
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 03/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:35

SUBJECT'S LAST NAME: GRAY

DATE/TIME OF TEST: 03/19/2016 05:02:07

Diagnostics Check	OK	04:55
Air Blank	0.000	04:56
Control Test	0.079	04:56
Air Blank	0.000	04:57
Subject Sample # 1	0.128	04:57
Breath Volume	2.019	04:57
Air Blank	0.000	04:58
Air Blank	0.000	05:00
Subject Sample # 2	0.117	05:00
Breath Volume	1.500	05:00
Air Blank	0.000	05:01
Control Test	0.079	05:01
Air Blank	0.000	05:02
Diagnostics Check	OK	05:02

Data Download Date/Time: 03/31/2016 12:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORTAGUS,B,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FISSEL,M
ARRESTING OFFICER AGENCY: Atlantic Beach PD
LAST AGENCY INSPECTION: 03/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 16:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:05

SUBJECT'S LAST NAME: EWING

DATE/TIME OF TEST: 03/19/2016 18:33:07

Diagnostics Check	OK	18:25
Air Blank	0.000	18:26
Control Test	0.078	18:26
Air Blank	0.000	18:26
Subject Sample # 1	0.136	18:28
Breath Volume	2.183	18:28
Air Blank	0.000	18:29
Air Blank	0.000	18:31
Subject Sample # 2	0.128	18:31
Breath Volume	1.949	18:31
Air Blank	0.000	18:32
Control Test	0.078	18:32
Air Blank	0.000	18:33
Diagnostics Check	OK	18:33

Data Download Date/Time: 03/31/2016 12:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCOTT,J,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MONTGOMERY,K
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 03/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:28

SUBJECT'S LAST NAME: BRISTOW

DATE/TIME OF TEST: 03/20/2016 04:56:26

Diagnostics Check	OK	04:49
Air Blank	0.000	04:49
Control Test	0.078	04:50
Air Blank	0.000	04:50
Subject Sample # 1	0.165	04:51
Breath Volume	1.765	04:51
Air Blank	0.000	04:52
Air Blank	0.000	04:53
Subject Sample # 2	0.170	04:54
Breath Volume	1.191	04:54
Air Blank	0.000	04:55
Control Test	0.078	04:55
Air Blank	0.000	04:56
Diagnostics Check	OK	04:56

Data Download Date/Time: 03/31/2016 12:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON, BRADEN, J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JUNG, B
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 19:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:03

SUBJECT'S LAST NAME: WASHINGTON

DATE/TIME OF TEST: 03/21/2016 01:29:58

Diagnostics Check	OK	01:24
Air Blank	0.000	01:24
Control Test	0.079	01:24
Air Blank	0.000	01:25
Subject Sample # 1	0.000	01:25
Breath Volume	1.621	01:25
Air Blank	0.000	01:26
Air Blank	0.000	01:28
Subject Sample # 2	0.000	01:28
Breath Volume	1.250	01:28
Air Blank	0.000	01:29
Control Test	0.079	01:29
Air Blank	0.000	01:29
Diagnostics Check	OK	01:29

Data Download Date/Time: 03/31/2016 12:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON, BRADEN, J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: EANES, J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:24

SUBJECT'S LAST NAME: ARCE-TOLEDO

DATE/TIME OF TEST: 03/21/2016 03:48:03

Diagnostics Check	OK	03:45	
Air Blank	0.000	03:45	
Control Test	0.080	03:46	
Air Blank	0.000	03:46	
Subject Sample	# 1	03:46	Subject Test Refused
Breath Volume	0.000	03:46	
Air Blank	0.000	03:47	
Control Test	0.080	03:47	
Air Blank	0.000	03:47	
Diagnostics Check	OK	03:48	

Data Download Date/Time: 03/31/2016 12:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLARIDA,K,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BORGERT,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 07:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:15

SUBJECT'S LAST NAME: REYES

DATE/TIME OF TEST: 03/21/2016 09:38:03

Diagnostics Check	OK	09:35	
Air Blank	0.000	09:35	
Control Test	0.081	09:36	
Air Blank	0.000	09:36	
Subject Sample	# 1	09:36	Subject Test Refused
Breath Volume	0.000	09:36	
Air Blank	0.000	09:37	
Control Test	0.081	09:37	
Air Blank	0.000	09:37	
Diagnostics Check	OK	09:38	

Data Download Date/Time: 03/31/2016 12:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTELL,CHRIS,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BYRNE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 16:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:07

SUBJECT'S LAST NAME: BORYS

DATE/TIME OF TEST: 03/21/2016 18:35:34

Diagnostics Check	OK	18:28
Air Blank	0.000	18:28
Control Test	0.081	18:28
Air Blank	0.000	18:29
Subject Sample # 1	0.000	18:30
Breath Volume	1.687	18:30
Air Blank	0.000	18:31
Air Blank	0.000	18:33
Subject Sample # 2	0.000	18:34
Breath Volume	1.722	18:34
Air Blank	0.000	18:34
Control Test	0.080	18:34
Air Blank	0.000	18:35
Diagnostics Check	OK	18:35

Data Download Date/Time: 03/31/2016 12:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORTAGUS, BLAKE, A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BAIS
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 17:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:45

SUBJECT'S LAST NAME: MOSS

DATE/TIME OF TEST: 03/22/2016 19:11:56

Diagnostics Check	OK	19:05
Air Blank	0.000	19:05
Control Test	0.080	19:06
Air Blank	0.000	19:06
Subject Sample # 1	0.000	19:07
Breath Volume	3.625	19:07
Air Blank	0.000	19:07
Air Blank	0.000	19:09
Subject Sample # 2	0.000	19:10
Breath Volume	2.117	19:10
Air Blank	0.000	19:11
Control Test	0.080	19:11
Air Blank	0.000	19:11
Diagnostics Check	OK	19:11

Data Download Date/Time: 03/31/2016 12:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORTAGUS, BLAKE, A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BROWN, T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 18:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:46

SUBJECT'S LAST NAME: CREWS

DATE/TIME OF TEST: 03/22/2016 20:19:58

Diagnostics Check	OK	20:07	
Air Blank	0.000	20:07	
Control Test	0.080	20:08	
Air Blank	0.000	20:08	
Subject Sample	# 1 0.000	20:09	
Breath Volume	2.054	20:09	
Air Blank	0.000	20:09	
Air Blank	0.000	20:11	
Subject Sample	# 2 0.000	20:15	Volume Not Met
Breath Volume	0.757	20:15	
Air Blank	0.000	20:15	
Air Blank	0.000	20:17	
Subject Sample	# 3 0.000	20:18	
Breath Volume	1.949	20:18	
Air Blank	0.000	20:19	
Control Test	0.079	20:19	
Air Blank	0.000	20:19	
Diagnostics Check	OK	20:19	

Data Download Date/Time: 03/31/2016 12:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALLIDAY,ROBERT,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: EANES,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 03/24/2016 00:32:02

Diagnostics Check	OK	00:25
Air Blank	0.000	00:26
Control Test	0.079	00:26
Air Blank	0.000	00:26
Subject Sample # 1	0.203	00:27
Breath Volume	1.660	00:27
Air Blank	0.000	00:27
Air Blank	0.000	00:29
Subject Sample # 2	0.192	00:30
Breath Volume	1.890	00:30
Air Blank	0.000	00:31
Control Test	0.079	00:31
Air Blank	0.000	00:31
Diagnostics Check	OK	00:32

Data Download Date/Time: 03/31/2016 12:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALLIDAY,ROBERT,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: REGISTER,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: WELDON

DATE/TIME OF TEST: 03/24/2016 02:14:04

Diagnostics Check	OK	02:11	
Air Blank	0.000	02:11	
Control Test	0.079	02:12	
Air Blank	0.000	02:12	
Subject Sample	# 1	02:12	Subject Test Refused
Breath Volume	0.000	02:12	
Air Blank	0.000	02:13	
Control Test	0.079	02:13	
Air Blank	0.000	02:13	
Diagnostics Check	OK	02:14	

Data Download Date/Time: 03/31/2016 12:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALLIDAY,ROBERT,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: GARCIA,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: LONGMEIER

DATE/TIME OF TEST: 03/24/2016 23:57:05

Diagnostics Check	OK	23:50
Air Blank	0.000	23:51
Control Test	0.079	23:51
Air Blank	0.000	23:52
Subject Sample # 1	0.000	23:52
Breath Volume	2.707	23:52
Air Blank	0.000	23:53
Air Blank	0.000	23:55
Subject Sample # 2	0.000	23:55
Breath Volume	2.320	23:55
Air Blank	0.000	23:56
Control Test	0.078	23:56
Air Blank	0.000	23:56
Diagnostics Check	OK	23:57

Data Download Date/Time: 03/31/2016 12:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALLIDAY,ROBERT,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MONTGOMERY,K
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 03/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 03/25/2016 03:56:26

Diagnostics Check	OK	03:50
Air Blank	0.000	03:50
Control Test	0.079	03:50
Air Blank	0.000	03:51
Subject Sample # 1	0.134	03:51
Breath Volume	1.988	03:51
Air Blank	0.000	03:52
Air Blank	0.000	03:54
Subject Sample # 2	0.139	03:54
Breath Volume	2.441	03:54
Air Blank	0.000	03:55
Control Test	0.079	03:55
Air Blank	0.000	03:56
Diagnostics Check	OK	03:56

Data Download Date/Time: 03/31/2016 12:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALLIDAY,ROBERT,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: TERRELL,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:31

SUBJECT'S LAST NAME: ETHIER-FINK

DATE/TIME OF TEST: 03/25/2016 23:57:54

Diagnostics Check	OK	23:51
Air Blank	0.000	23:52
Control Test	0.079	23:52
Air Blank	0.000	23:53
Subject Sample # 1	0.171	23:53
Breath Volume	2.484	23:53
Air Blank	0.000	23:54
Air Blank	0.000	23:56
Subject Sample # 2	0.172	23:56
Breath Volume	2.160	23:56
Air Blank	0.000	23:56
Control Test	0.078	23:57
Air Blank	0.000	23:57
Diagnostics Check	OK	23:57

Data Download Date/Time: 03/31/2016 12:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALLIDAY,ROBERT,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: GARCIA,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 03/26/2016 03:16:24

Diagnostics Check	OK	03:10
Air Blank	0.000	03:10
Control Test	0.078	03:11
Air Blank	0.000	03:11
Subject Sample # 1	0.234	03:12
Breath Volume	2.816	03:12
Air Blank	0.000	03:12
Air Blank	0.000	03:14
Subject Sample # 2	0.237	03:14
Breath Volume	2.808	03:14
Air Blank	0.000	03:15
Control Test	0.078	03:15
Air Blank	0.000	03:16
Diagnostics Check	OK	03:16

Data Download Date/Time: 03/31/2016 12:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALLIDAY,ROBERT,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: TERRELL,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:47

SUBJECT'S LAST NAME: LAM

DATE/TIME OF TEST: 03/26/2016 04:10:01

Diagnostics Check	OK	04:07	
Air Blank	0.000	04:07	
Control Test	0.078	04:07	
Air Blank	0.000	04:08	
Subject Sample	# 1	04:08	Subject Test Refused
Breath Volume	0.000	04:08	
Air Blank	0.000	04:09	
Control Test	0.078	04:09	
Air Blank	0.000	04:09	
Diagnostics Check	OK	04:10	

Data Download Date/Time: 03/31/2016 12:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAKER, TYLER, R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: KEEN, C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 17:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:20

SUBJECT'S LAST NAME: LANIER

DATE/TIME OF TEST: 03/26/2016 19:47:42

Diagnostics Check	OK	19:41
Air Blank	0.000	19:41
Control Test	0.079	19:41
Air Blank	0.000	19:42
Subject Sample # 1	0.055	19:43
Breath Volume	2.914	19:43
Air Blank	0.000	19:43
Air Blank	0.000	19:45
Subject Sample # 2	0.052	19:46
Breath Volume	2.515	19:46
Air Blank	0.000	19:46
Control Test	0.078	19:47
Air Blank	0.000	19:47
Diagnostics Check	OK	19:47

Data Download Date/Time: 03/31/2016 12:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALLIDAY,ROBERT,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DOYLE,B
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 03/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:28

SUBJECT'S LAST NAME: GOLDBACH

DATE/TIME OF TEST: 03/27/2016 22:50:56

Diagnostics Check	OK	22:48	
Air Blank	0.000	22:48	
Control Test	0.080	22:48	
Air Blank	0.000	22:49	
Subject Sample	# 1	22:49	Subject Test Refused
Breath Volume	0.000	22:49	
Air Blank	0.000	22:50	
Control Test	0.079	22:50	
Air Blank	0.000	22:50	
Diagnostics Check	OK	22:50	

Data Download Date/Time: 03/31/2016 12:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALLIDAY,ROBERT,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JOHNSTON,C
ARRESTING OFFICER AGENCY: Atlantic Beach PD
LAST AGENCY INSPECTION: 03/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:07

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 03/28/2016 04:29:58

Diagnostics Check	OK	04:27	
Air Blank	0.000	04:27	
Control Test	0.080	04:27	
Air Blank	0.000	04:28	
Subject Sample	# 1	04:28	Subject Test Refused
Breath Volume	0.000	04:28	
Air Blank	0.000	04:29	
Control Test	0.080	04:29	
Air Blank	0.000	04:29	
Diagnostics Check	OK	04:29	

Data Download Date/Time: 03/31/2016 12:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON, BRADEN, J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BAIS
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:54

SUBJECT'S LAST NAME: RADER

DATE/TIME OF TEST: 03/30/2016 02:17:17

Diagnostics Check	OK	02:14	
Air Blank	0.000	02:14	
Control Test	0.079	02:15	
Air Blank	0.000	02:15	
Subject Sample	# 1	02:15	Subject Test Refused
Breath Volume	0.000	02:15	
Air Blank	0.000	02:16	
Control Test	0.079	02:16	
Air Blank	0.000	02:17	
Diagnostics Check	OK	02:17	

Data Download Date/Time: 03/31/2016 12:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON, BRADEN, J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BURNETT, B
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:32

SUBJECT'S LAST NAME: MEHBOOB

DATE/TIME OF TEST: 03/30/2016 04:55:10

Diagnostics Check	OK	04:52	
Air Blank	0.000	04:52	
Control Test	0.080	04:53	
Air Blank	0.000	04:53	
Subject Sample	# 1	04:53	Subject Test Refused
Breath Volume	0.000	04:53	
Air Blank	0.000	04:54	
Control Test	0.079	04:54	
Air Blank	0.000	04:55	
Diagnostics Check	OK	04:55	

Data Download Date/Time: 03/31/2016 12:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON, BRADEN, J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BRYANT, R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:09

SUBJECT'S LAST NAME: MILLEDGE

DATE/TIME OF TEST: 03/31/2016 02:35:54

Diagnostics Check	OK	02:29
Air Blank	0.000	02:30
Control Test	0.079	02:30
Air Blank	0.000	02:30
Subject Sample # 1	0.111	02:31
Breath Volume	1.324	02:31
Air Blank	0.000	02:32
Air Blank	0.000	02:33
Subject Sample # 2	0.108	02:34
Breath Volume	1.621	02:34
Air Blank	0.000	02:35
Control Test	0.079	02:35
Air Blank	0.000	02:35
Diagnostics Check	OK	02:35

Data Download Date/Time: 03/31/2016 12:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLARIDA,K,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: PARSLEY,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 05:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:30

SUBJECT'S LAST NAME: HAMILTON

DATE/TIME OF TEST: 03/31/2016 07:56:07

Diagnostics Check	OK	07:53	
Air Blank	0.000	07:53	
Control Test	0.079	07:54	
Air Blank	0.000	07:54	
Subject Sample	# 1	07:54	Subject Test Refused
Breath Volume	0.000	07:54	
Air Blank	0.000	07:55	
Control Test	0.079	07:55	
Air Blank	0.000	07:56	
Diagnostics Check	OK	07:56	

Data Download Date/Time: 03/31/2016 12:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,J,R
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: JONES,D
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 02/16/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A3
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 11:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 11:58

SUBJECT'S LAST NAME: REESE

DATE/TIME OF TEST: 03/04/2016 12:24:34

Diagnostics Check	OK	12:18
Air Blank	0.000	12:18
Control Test	0.081	12:19
Air Blank	0.000	12:19
Subject Sample # 1	0.207	12:20
Breath Volume	2.542	12:20
Air Blank	0.000	12:20
Air Blank	0.000	12:22
Subject Sample # 2	0.204	12:23
Breath Volume	2.027	12:23
Air Blank	0.000	12:23
Control Test	0.081	12:23
Air Blank	0.000	12:24
Diagnostics Check	OK	12:24

Data Download Date/Time: 03/28/2016 07:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLIMON,SHANIESA,L
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: MCNEIL,S
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 02/16/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A3
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:07

SUBJECT'S LAST NAME: KEEN

DATE/TIME OF TEST: 03/05/2016 03:35:01

Diagnostics Check	OK	03:27
Air Blank	0.000	03:28
Control Test	0.081	03:28
Air Blank	0.000	03:29
Subject Sample # 1	0.154	03:30
Breath Volume	3.007	03:30
Air Blank	0.000	03:30
Air Blank	0.000	03:32
Subject Sample # 2	0.149	03:33
Breath Volume	2.519	03:33
Air Blank	0.000	03:34
Control Test	0.079	03:34
Air Blank	0.000	03:34
Diagnostics Check	OK	03:35

Data Download Date/Time: 03/28/2016 07:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HEALY,G,B
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: HEALU,G
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 02/16/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A3
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:49

SUBJECT'S LAST NAME: MCMULLEN

DATE/TIME OF TEST: 03/06/2016 03:15:42

Diagnostics Check	OK	03:09
Air Blank	0.000	03:09
Control Test	0.081	03:10
Air Blank	0.000	03:10
Subject Sample # 1	0.217	03:11
Breath Volume	4.585	03:11
Air Blank	0.000	03:11
Air Blank	0.000	03:13
Subject Sample # 2	0.209	03:14
Breath Volume	3.335	03:14
Air Blank	0.000	03:14
Control Test	0.081	03:15
Air Blank	0.000	03:15
Diagnostics Check	OK	03:15

Data Download Date/Time: 03/28/2016 07:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HEALY,G,B
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: HEALY,G
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 02/16/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A3
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:34

SUBJECT'S LAST NAME: LOPES

DATE/TIME OF TEST: 03/07/2016 02:05:08

Diagnostics Check	OK	01:55
Air Blank	0.000	01:55
Control Test	0.081	01:55
Air Blank	0.000	01:56
Subject Sample # 1	0.114	01:56
Breath Volume	1.132	01:56
Air Blank	0.000	01:57
Air Blank	0.000	01:59
Subject Sample # 2	0.140	02:00
Breath Volume	3.992	02:00
Air Blank	0.000	02:01
Air Blank	0.000	02:03
Subject Sample # 3	0.117	02:03
Breath Volume	1.378	02:03
Air Blank	0.000	02:04
Control Test	0.080	02:04
Air Blank	0.000	02:05
Diagnostics Check	OK	02:05

Data Download Date/Time: 03/28/2016 07:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HEALY,G,B
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: HEALY,G
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 02/16/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A3
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:21

SUBJECT'S LAST NAME: BACH

DATE/TIME OF TEST: 03/10/2016 23:47:52

Diagnostics Check	OK	23:41
Air Blank	0.000	23:42
Control Test	0.081	23:42
Air Blank	0.000	23:43
Subject Sample # 1	0.148	23:43
Breath Volume	1.750	23:43
Air Blank	0.000	23:44
Air Blank	0.000	23:45
Subject Sample # 2	0.156	23:46
Breath Volume	2.472	23:46
Air Blank	0.000	23:46
Control Test	0.080	23:47
Air Blank	0.000	23:47
Diagnostics Check	OK	23:47

Data Download Date/Time: 03/28/2016 07:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIEDSMA, TERRY, D
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: PATTERSON, W
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 02/16/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A3
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: SIEDSMA

DATE/TIME OF TEST: 03/12/2016 00:25:33

Diagnostics Check	OK	00:22	
Air Blank	0.000	00:23	
Control Test	0.081	00:23	
Air Blank	0.000	00:23	
Subject Sample	# 1	00:24	Subject Test Refused
Breath Volume	0.000	00:24	
Air Blank	0.000	00:24	
Control Test	0.080	00:24	
Air Blank	0.000	00:25	
Diagnostics Check	OK	00:25	

Data Download Date/Time: 03/28/2016 07:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HODGES,POLLY,E
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: HODGES,652
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 02/16/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A3
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:00

SUBJECT'S LAST NAME: PHELPS

DATE/TIME OF TEST: 03/12/2016 06:35:05

Diagnostics Check	OK	06:28
Air Blank	0.000	06:28
Control Test	0.081	06:29
Air Blank	0.000	06:29
Subject Sample # 1	0.105	06:30
Breath Volume	1.734	06:30
Air Blank	0.000	06:30
Air Blank	0.000	06:32
Subject Sample # 2	0.111	06:33
Breath Volume	1.460	06:33
Air Blank	0.000	06:34
Control Test	0.080	06:34
Air Blank	0.000	06:35
Diagnostics Check	OK	06:35

Data Download Date/Time: 03/28/2016 07:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HEALY,G,B
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: HEALY,G
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 02/16/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A3
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:15

SUBJECT'S LAST NAME: JOINER

DATE/TIME OF TEST: 03/12/2016 21:41:23

Diagnostics Check	OK	21:35
Air Blank	0.000	21:35
Control Test	0.080	21:36
Air Blank	0.000	21:36
Subject Sample # 1	0.026	21:37
Breath Volume	1.460	21:37
Air Blank	0.000	21:37
Air Blank	0.000	21:39
Subject Sample # 2	0.027	21:39
Breath Volume	1.605	21:39
Air Blank	0.000	21:40
Control Test	0.079	21:40
Air Blank	0.000	21:41
Diagnostics Check	OK	21:41

Data Download Date/Time: 03/28/2016 07:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HEALY,G,B
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: HEALY,G
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 02/16/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A3
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: DWINNELL

DATE/TIME OF TEST: 03/13/2016 03:37:50

Diagnostics Check		OK	03:31
Air Blank		0.000	03:32
Control Test		0.081	03:32
Air Blank		0.000	03:32
Subject Sample	# 1	0.139	03:33
Breath Volume		1.828	03:33
Air Blank		0.000	03:33
Air Blank		0.000	03:35
Subject Sample	# 2	0.145	03:36
Breath Volume		1.996	03:36
Air Blank		0.000	03:36
Control Test		0.080	03:37
Air Blank		0.000	03:37
Diagnostics Check		OK	03:37

Data Download Date/Time: 03/28/2016 07:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCNEIL,K,P
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: DOUGLAS,D
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 02/16/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A3
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:28

SUBJECT'S LAST NAME: MANNING

DATE/TIME OF TEST: 03/14/2016 21:55:39

Diagnostics Check	OK	21:48
Air Blank	0.000	21:49
Control Test	0.080	21:49
Air Blank	0.000	21:49
Subject Sample # 1	0.123	21:50
Breath Volume	1.222	21:50
Air Blank	0.000	21:51
Air Blank	0.000	21:53
Subject Sample # 2	0.119	21:54
Breath Volume	1.457	21:54
Air Blank	0.000	21:54
Control Test	0.079	21:55
Air Blank	0.000	21:55
Diagnostics Check	OK	21:55

Data Download Date/Time: 03/28/2016 07:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHOOK,J,S
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: DELACRUZ,M
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 02/16/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A3
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: JOHNS

DATE/TIME OF TEST: 03/15/2016 22:50:59

Diagnostics Check		OK	22:37	
Air Blank		0.000	22:37	
Control Test		0.080	22:38	
Air Blank		0.000	22:38	
Subject Sample	# 1	0.141	22:41	Volume Not Met
Breath Volume		0.125	22:41	
Air Blank		0.000	22:42	
Air Blank		0.000	22:44	
Subject Sample	# 2	0.146	22:45	
Breath Volume		1.445	22:45	
Air Blank		0.000	22:45	
Air Blank		0.000	22:47	
Subject Sample	# 3	0.138	22:49	
Breath Volume		1.394	22:49	
Air Blank		0.000	22:50	
Control Test		0.078	22:50	
Air Blank		0.000	22:50	
Diagnostics Check		OK	22:50	

Data Download Date/Time: 03/28/2016 07:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HEALY,G,B
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: HEALY,G
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 02/16/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A3
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:08

SUBJECT'S LAST NAME: JANSEN

DATE/TIME OF TEST: 03/19/2016 02:35:08

Diagnostics Check	OK	02:28
Air Blank	0.000	02:29
Control Test	0.081	02:29
Air Blank	0.000	02:30
Subject Sample # 1	0.170	02:30
Breath Volume	3.015	02:30
Air Blank	0.000	02:31
Air Blank	0.000	02:33
Subject Sample # 2	0.157	02:33
Breath Volume	1.386	02:33
Air Blank	0.000	02:34
Control Test	0.080	02:34
Air Blank	0.000	02:35
Diagnostics Check	OK	02:35

Data Download Date/Time: 03/28/2016 07:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HEALY,G,B
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: HEALY,G
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 02/16/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A3
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 19:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:26

SUBJECT'S LAST NAME: WERNET

DATE/TIME OF TEST: 03/20/2016 20:53:00

Diagnostics Check	OK	20:46
Air Blank	0.000	20:47
Control Test	0.080	20:47
Air Blank	0.000	20:48
Subject Sample # 1	0.084	20:48
Breath Volume	3.480	20:48
Air Blank	0.000	20:49
Air Blank	0.000	20:50
Subject Sample # 2	0.084	20:51
Breath Volume	4.144	20:51
Air Blank	0.000	20:52
Control Test	0.080	20:52
Air Blank	0.000	20:52
Diagnostics Check	OK	20:53

Data Download Date/Time: 03/28/2016 07:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLIMON,SHANIESA,L
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: MCNEIL,SCOTT
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 02/16/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A3
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:10

SUBJECT'S LAST NAME: DORCHNIC

DATE/TIME OF TEST: 03/20/2016 21:39:50

Diagnostics Check	OK	21:33
Air Blank	0.000	21:33
Control Test	0.078	21:34
Air Blank	0.000	21:34
Subject Sample # 1	0.195	21:35
Breath Volume	3.640	21:35
Air Blank	0.000	21:35
Air Blank	0.000	21:37
Subject Sample # 2	0.199	21:38
Breath Volume	3.054	21:38
Air Blank	0.000	21:38
Control Test	0.080	21:39
Air Blank	0.000	21:39
Diagnostics Check	OK	21:39

Data Download Date/Time: 03/28/2016 07:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPBELL,D,R
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: DELEON,L
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 02/16/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A3
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:27

SUBJECT'S LAST NAME: CAMACHO

DATE/TIME OF TEST: 03/26/2016 01:56:01

Diagnostics Check		OK	01:49
Air Blank		0.000	01:49
Control Test		0.080	01:49
Air Blank		0.000	01:50
Subject Sample	# 1	0.206	01:51
Breath Volume		2.667	01:51
Air Blank		0.000	01:51
Air Blank		0.000	01:53
Subject Sample	# 2	0.203	01:54
Breath Volume		2.277	01:54
Air Blank		0.000	01:55
Control Test		0.079	01:55
Air Blank		0.000	01:55
Diagnostics Check		OK	01:56

Data Download Date/Time: 03/28/2016 07:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GORDON,GREG,M
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: DELEON,LIZANDRO
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 02/16/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A3
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:17

SUBJECT'S LAST NAME: WASHINGTON

DATE/TIME OF TEST: 03/26/2016 23:41:00

Diagnostics Check	OK	23:38	
Air Blank	0.000	23:38	
Control Test	0.082	23:38	
Air Blank	0.000	23:39	
Subject Sample	# 1	23:39	Subject Test Refused
Breath Volume	0.000	23:39	
Air Blank	0.000	23:40	
Control Test	0.082	23:40	
Air Blank	0.000	23:40	
Diagnostics Check	OK	23:41	

Data Download Date/Time: 03/28/2016 07:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STOCKLING,T,N
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: DELEON,L
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 02/16/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A3
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:01

SUBJECT'S LAST NAME: GRIFFIN

DATE/TIME OF TEST: 03/27/2016 03:25:17

Diagnostics Check	OK	03:22	
Air Blank	0.000	03:22	
Control Test	0.082	03:23	
Air Blank	0.000	03:23	
Subject Sample	# 1	03:23	Subject Test Refused
Breath Volume	0.000	03:23	
Air Blank	0.000	03:24	
Control Test	0.081	03:24	
Air Blank	0.000	03:25	
Diagnostics Check	OK	03:25	

Data Download Date/Time: 03/28/2016 07:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001282

INSTRUMENT REGISTERED TO: UNIVERSITY OF FL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANDERSON,GEOFFREY,A
BREATH TEST OPERATOR AGENCY: U of FI PD
ARRESTING OFFICER: BAIN,KENNETH
ARRESTING OFFICER AGENCY: U of FI PD
LAST AGENCY INSPECTION: 02/18/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07014080A2
DRY GAS EXPIRY DATE: 05/01/2016
VIOLATION CODE: DUI
ARREST TIME: 01:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:14

SUBJECT'S LAST NAME: GILMORE

DATE/TIME OF TEST: 03/02/2016 02:40:59

Diagnostics Check	OK	02:34	
Air Blank	0.000	02:35	
Control Test	0.077	02:35	
Air Blank	0.000	02:36	
Subject Sample	# 1 0.105	02:36	
Breath Volume	1.425	02:36	
Air Blank	0.000	02:37	
Air Blank	0.000	02:39	
Subject Sample	# 2	02:39	Subject Test Refused
Breath Volume	0.000	02:39	
Air Blank	0.000	02:40	
Control Test	0.076	02:40	
Air Blank	0.000	02:40	
Diagnostics Check	OK	02:40	

Data Download Date/Time: 03/29/2016 14:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001282

INSTRUMENT REGISTERED TO: UNIVERSITY OF FL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANDERSON,GEOFFREY,A
BREATH TEST OPERATOR AGENCY: U of FI PD
ARRESTING OFFICER: BAIN,KENNETH
ARRESTING OFFICER AGENCY: U of FI PD
LAST AGENCY INSPECTION: 02/18/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07014080A2
DRY GAS EXPIRY DATE: 05/01/2016
VIOLATION CODE: DUI
ARREST TIME: 02:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:42

SUBJECT'S LAST NAME: MOSLEY

DATE/TIME OF TEST: 03/05/2016 03:08:56

Diagnostics Check		OK	03:02
Air Blank		0.000	03:03
Control Test		0.077	03:03
Air Blank		0.000	03:04
Subject Sample	# 1	0.128	03:04
Breath Volume		1.421	03:04
Air Blank		0.000	03:05
Air Blank		0.000	03:06
Subject Sample	# 2	0.129	03:07
Breath Volume		2.609	03:07
Air Blank		0.000	03:08
Control Test		0.078	03:08
Air Blank		0.000	03:08
Diagnostics Check		OK	03:08

Data Download Date/Time: 03/29/2016 14:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001282

INSTRUMENT REGISTERED TO: UNIVERSITY OF FL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCORMICK,ERIC,R
BREATH TEST OPERATOR AGENCY: U of FI PD
ARRESTING OFFICER: HENDERSON,TREVER
ARRESTING OFFICER AGENCY: U of FI PD
LAST AGENCY INSPECTION: 02/18/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07014080A2
DRY GAS EXPIRY DATE: 05/01/2016
VIOLATION CODE: DUI
ARREST TIME: 00:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:33

SUBJECT'S LAST NAME: CAHILL

DATE/TIME OF TEST: 03/08/2016 01:07:31

Diagnostics Check	OK	00:54	
Air Blank	0.000	00:55	
Control Test	0.078	00:55	
Air Blank	0.000	00:56	
Subject Sample # 1	0.251	00:56	
Breath Volume	1.117	00:56	
Air Blank	0.000	00:57	
Air Blank	0.000	00:59	
Subject Sample # 2	0.216	01:02	Volume Not Met
Breath Volume	0.714	01:02	
Air Blank	0.000	01:03	
Air Blank	0.000	01:05	
Subject Sample # 3	0.239	01:05	
Breath Volume	1.382	01:05	
Air Blank	0.000	01:06	
Control Test	0.073	01:06	Control Outside Tolerance
Air Blank	0.000	01:07	

Data Download Date/Time: 03/29/2016 14:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001282

INSTRUMENT REGISTERED TO: UNIVERSITY OF FL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCORMICK,ERIC,R
BREATH TEST OPERATOR AGENCY: U of FI PD
ARRESTING OFFICER: HENDERSON,TREVER
ARRESTING OFFICER AGENCY: U of FI PD
LAST AGENCY INSPECTION: 02/18/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07014080A2
DRY GAS EXPIRY DATE: 05/01/2016
VIOLATION CODE: DUI
ARREST TIME: 00:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:33

SUBJECT'S LAST NAME: CAHILL

DATE/TIME OF TEST: 03/08/2016 01:21:47

Diagnostics Check		OK	01:13
Air Blank		0.000	01:14
Control Test		0.076	01:14
Air Blank		0.000	01:15
Subject Sample	# 1	0.247	01:17
Breath Volume		1.406	01:17
Air Blank		0.000	01:17
Air Blank		0.000	01:19
Subject Sample	# 2	0.241	01:20
Breath Volume		1.296	01:20
Air Blank		0.000	01:20
Control Test		0.075	01:21
Air Blank		0.000	01:21
Diagnostics Check		OK	01:21

Data Download Date/Time: 03/29/2016 14:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001282

INSTRUMENT REGISTERED TO: UNIVERSITY OF FL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUNT,CHARLES,R
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: LARA,ALEX
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 02/18/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07014080A2
DRY GAS EXPIRY DATE: 05/01/2016
VIOLATION CODE: DUI
ARREST TIME: 00:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: WILKINSON

DATE/TIME OF TEST: 03/09/2016 01:23:33

Diagnostics Check	OK	01:12	
Air Blank	0.000	01:12	
Control Test	0.078	01:12	
Air Blank	0.000	01:13	
Subject Sample	# 1	01:16	No Sample Provided
Breath Volume	0.000	01:16	
Air Blank	0.000	01:17	
Air Blank	0.000	01:19	
Subject Sample	# 2	01:22	No Sample Provided
Breath Volume	0.000	01:22	
Air Blank	0.000	01:22	
Control Test	0.079	01:22	
Air Blank	0.000	01:23	
Diagnostics Check	OK	01:23	

Data Download Date/Time: 03/29/2016 14:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001282

INSTRUMENT REGISTERED TO: UNIVERSITY OF FL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORGAN,RICKY,L
BREATH TEST OPERATOR AGENCY: U of FI PD
ARRESTING OFFICER: NGUYEN,RICK
ARRESTING OFFICER AGENCY: U of FI PD
LAST AGENCY INSPECTION: 02/18/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07014080A2
DRY GAS EXPIRY DATE: 05/01/2016
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:44

SUBJECT'S LAST NAME: HASANOV

DATE/TIME OF TEST: 03/12/2016 03:29:04

Diagnostics Check	OK	03:20	
Air Blank	0.000	03:20	
Control Test	0.078	03:21	
Air Blank	0.000	03:21	
Subject Sample	# 1	03:24	No Sample Provided
Breath Volume	0.000	03:24	
Air Blank	0.000	03:25	
Air Blank	0.000	03:27	
Subject Sample	# 2	03:27	Subject Test Refused
Breath Volume	0.000	03:27	
Air Blank	0.000	03:28	
Control Test	0.079	03:28	
Air Blank	0.000	03:28	
Diagnostics Check	OK	03:29	

Data Download Date/Time: 03/29/2016 14:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001282

INSTRUMENT REGISTERED TO: UNIVERSITY OF FL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PETERSON,KYLE,A
BREATH TEST OPERATOR AGENCY: U of FI PD
ARRESTING OFFICER: MCMILLAN,SHIKINA
ARRESTING OFFICER AGENCY: U of FI PD
LAST AGENCY INSPECTION: 02/18/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07014080A2
DRY GAS EXPIRY DATE: 05/01/2016
VIOLATION CODE: DUI
ARREST TIME: 04:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 03/13/2016 04:50:06

Diagnostics Check	OK	04:45	
Air Blank	0.000	04:45	
Control Test	0.078	04:45	
Air Blank	0.000	04:46	
Subject Sample	# 1	04:48	Subject Test Refused
Breath Volume	0.000	04:48	
Air Blank	0.000	04:49	
Control Test	0.079	04:49	
Air Blank	0.000	04:50	
Diagnostics Check	OK	04:50	

Data Download Date/Time: 03/29/2016 14:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001282

INSTRUMENT REGISTERED TO: UNIVERSITY OF FL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PETERSON,KYLE,A
BREATH TEST OPERATOR AGENCY: U of FI PD
ARRESTING OFFICER: MCMILLAN,SHIKINA
ARRESTING OFFICER AGENCY: U of FI PD
LAST AGENCY INSPECTION: 02/18/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07014080A2
DRY GAS EXPIRY DATE: 05/01/2016
VIOLATION CODE: DUI
ARREST TIME: 01:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:33

SUBJECT'S LAST NAME: HOLMES

DATE/TIME OF TEST: 03/18/2016 02:11:17

Diagnostics Check		OK	01:55	
Air Blank		0.000	01:56	
Control Test		0.078	01:56	
Air Blank		0.000	01:57	
Subject Sample	# 1	0.174	02:00	Volume Not Met
Breath Volume		0.082	02:00	
Air Blank		0.000	02:01	
Air Blank		0.000	02:02	
Subject Sample	# 2	0.214	02:03	
Breath Volume		1.351	02:03	
Air Blank		0.000	02:04	
Air Blank		0.000	02:06	
Subject Sample	# 3	0.210	02:09	Volume Not Met
Breath Volume		0.644	02:09	
Air Blank		0.000	02:10	
Control Test		0.075	02:10	
Air Blank		0.000	02:11	
Diagnostics Check		OK	02:11	

Data Download Date/Time: 03/29/2016 14:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001282

INSTRUMENT REGISTERED TO: UNIVERSITY OF FL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PETERSON,KYLE,A
BREATH TEST OPERATOR AGENCY: U of FI PD
ARRESTING OFFICER: MCMILLAN,SHIKINA
ARRESTING OFFICER AGENCY: U of FI PD
LAST AGENCY INSPECTION: 02/18/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07014080A2
DRY GAS EXPIRY DATE: 05/01/2016
VIOLATION CODE: DUI
ARREST TIME: 01:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:33

SUBJECT'S LAST NAME: HOLMES

DATE/TIME OF TEST: 03/18/2016 02:29:29

Diagnostics Check	OK	02:13	
Air Blank	0.000	02:14	
Control Test	0.075	02:14	
Air Blank	0.000	02:15	
Subject Sample	# 1 0.191	02:17	
Breath Volume	1.617	02:17	
Air Blank	0.000	02:17	
Air Blank	0.000	02:19	
Subject Sample	# 2 0.178	02:22	Volume Not Met
Breath Volume	0.210	02:22	
Air Blank	0.000	02:23	
Air Blank	0.000	02:25	
Subject Sample	# 3 0.199	02:27	
Breath Volume	2.300	02:27	
Air Blank	0.000	02:28	
Control Test	0.073	02:28	Control Outside Tolerance
Air Blank	0.000	02:29	

Data Download Date/Time: 03/29/2016 14:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001282

INSTRUMENT REGISTERED TO: UNIVERSITY OF FL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PETERSON,KYLE,A
BREATH TEST OPERATOR AGENCY: U of FI PD
ARRESTING OFFICER: MCMILLAN,SHIKINA
ARRESTING OFFICER AGENCY: U of FI PD
LAST AGENCY INSPECTION: 02/18/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07014080A2
DRY GAS EXPIRY DATE: 05/01/2016
VIOLATION CODE: DUI
ARREST TIME: 01:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:33

SUBJECT'S LAST NAME: HOLMES

DATE/TIME OF TEST: 03/18/2016 02:43:38

Diagnostics Check	OK	02:40	
Air Blank	0.000	02:41	
Control Test	0.076	02:41	
Air Blank	0.000	02:41	
Subject Sample	# 1	02:42	Subject Test Refused
Breath Volume	0.000	02:42	
Air Blank	0.000	02:42	
Control Test	0.077	02:43	
Air Blank	0.000	02:43	
Diagnostics Check	OK	02:43	

Data Download Date/Time: 03/29/2016 14:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001282

INSTRUMENT REGISTERED TO: UNIVERSITY OF FL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCORMICK,ERIC,R
BREATH TEST OPERATOR AGENCY: U of FI PD
ARRESTING OFFICER: MCCORMICK,ERIC
ARRESTING OFFICER AGENCY: U of FI PD
LAST AGENCY INSPECTION: 02/18/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07014080A2
DRY GAS EXPIRY DATE: 05/01/2016
VIOLATION CODE: DUI
ARREST TIME: 02:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: STOUT

DATE/TIME OF TEST: 03/26/2016 04:03:04

Diagnostics Check	OK	03:56
Air Blank	0.000	03:57
Control Test	0.078	03:57
Air Blank	0.000	03:58
Subject Sample # 1	0.144	03:58
Breath Volume	2.535	03:58
Air Blank	0.000	03:59
Air Blank	0.000	04:01
Subject Sample # 2	0.145	04:01
Breath Volume	2.687	04:01
Air Blank	0.000	04:02
Control Test	0.077	04:02
Air Blank	0.000	04:02
Diagnostics Check	OK	04:03

Data Download Date/Time: 03/29/2016 14:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001285

INSTRUMENT REGISTERED TO: GILCHRIST COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUTLER,ADAM,S
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: BUTLER,ADAM
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 02/19/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 259585
DRY GAS EXPIRY DATE: 01/12/2019
VIOLATION CODE: DUI
ARREST TIME: 08:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:50

SUBJECT'S LAST NAME: ROBINSON

DATE/TIME OF TEST: 03/02/2016 09:56:24

Diagnostics Check	OK	09:50
Air Blank	0.000	09:50
Control Test	0.079	09:51
Air Blank	0.000	09:51
Subject Sample # 1	0.000	09:52
Breath Volume	3.085	09:52
Air Blank	0.000	09:52
Air Blank	0.000	09:54
Subject Sample # 2	0.000	09:55
Breath Volume	2.851	09:55
Air Blank	0.000	09:55
Control Test	0.080	09:55
Air Blank	0.000	09:56
Diagnostics Check	OK	09:56

Data Download Date/Time: 04/01/2016 23:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001285

INSTRUMENT REGISTERED TO: GILCHRIST COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUTLER,ADAM,S
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: BUTLER,ADAM
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 02/19/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 259585
DRY GAS EXPIRY DATE: 01/12/2019
VIOLATION CODE: DUI
ARREST TIME: 18:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:35

SUBJECT'S LAST NAME: SANCHEZ

DATE/TIME OF TEST: 03/30/2016 19:23:31

Diagnostics Check	OK	19:16
Air Blank	0.000	19:17
Control Test	0.078	19:17
Air Blank	0.000	19:17
Subject Sample # 1	0.209	19:19
Breath Volume	2.097	19:19
Air Blank	0.000	19:19
Air Blank	0.000	19:21
Subject Sample # 2	0.199	19:21
Breath Volume	1.945	19:21
Air Blank	0.000	19:22
Control Test	0.077	19:22
Air Blank	0.000	19:23
Diagnostics Check	OK	19:23

Data Download Date/Time: 04/01/2016 23:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001287

INSTRUMENT REGISTERED TO: BAKER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FROST,KEVIN,W
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: FROST,KEVIN
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 02/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A1
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: GREEN

DATE/TIME OF TEST: 03/10/2016 00:48:12

Diagnostics Check	OK	00:45	
Air Blank	0.000	00:45	
Control Test	0.082	00:46	
Air Blank	0.000	00:46	
Subject Sample	# 1	00:46	Subject Test Refused
Breath Volume	0.000	00:46	
Air Blank	0.000	00:47	
Control Test	0.082	00:47	
Air Blank	0.000	00:48	
Diagnostics Check	OK	00:48	

Data Download Date/Time: 03/29/2016 05:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001287

INSTRUMENT REGISTERED TO: BAKER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARLEY,JOSEPH,E
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: FARLEY,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 02/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A1
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 19:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:10

SUBJECT'S LAST NAME: MIER

DATE/TIME OF TEST: 03/26/2016 20:36:49

Diagnostics Check	OK	20:30
Air Blank	0.000	20:31
Control Test	0.082	20:31
Air Blank	0.000	20:31
Subject Sample # 1	0.084	20:32
Breath Volume	2.066	20:32
Air Blank	0.000	20:33
Air Blank	0.000	20:34
Subject Sample # 2	0.075	20:35
Breath Volume	1.398	20:35
Air Blank	0.000	20:35
Control Test	0.081	20:36
Air Blank	0.000	20:36
Diagnostics Check	OK	20:36

Data Download Date/Time: 03/29/2016 05:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001287

INSTRUMENT REGISTERED TO: BAKER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DRIGGERS,RODNEY,M
BREATH TEST OPERATOR AGENCY: Baker County SO
ARRESTING OFFICER: HEATH,MARC
ARRESTING OFFICER AGENCY: Baker County SO
LAST AGENCY INSPECTION: 03/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A1
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 03/31/2016 23:43:06

Diagnostics Check	OK	23:35
Air Blank	0.000	23:36
Control Test	0.082	23:36
Air Blank	0.000	23:36
Subject Sample # 1	0.238	23:38
Breath Volume	1.746	23:38
Air Blank	0.000	23:38
Air Blank	0.000	23:40
Subject Sample # 2	0.235	23:41
Breath Volume	1.832	23:41
Air Blank	0.000	23:42
Control Test	0.080	23:42
Air Blank	0.000	23:43
Diagnostics Check	OK	23:43

Data Download Date/Time: 04/30/2016 02:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001289

INSTRUMENT REGISTERED TO: UNION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TOWNSEND, CHARLES, W
BREATH TEST OPERATOR AGENCY: Union County SO
ARRESTING OFFICER: ANDREWS, SCOTT
ARRESTING OFFICER AGENCY: Union County SO
LAST AGENCY INSPECTION: 03/01/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1824329
DRY GAS EXPIRY DATE: 03/09/2018
VIOLATION CODE: ZERO TOL
ARREST TIME: 18:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:40

SUBJECT'S LAST NAME: POWELL

DATE/TIME OF TEST: 03/05/2016 19:18:29

Diagnostics Check	OK	19:12
Air Blank	0.000	19:12
Control Test	0.080	19:13
Air Blank	0.000	19:13
Subject Sample # 1	0.173	19:14
Breath Volume	1.589	19:14
Air Blank	0.000	19:14
Air Blank	0.000	19:16
Subject Sample # 2	0.174	19:16
Breath Volume	1.234	19:16
Air Blank	0.000	19:17
Control Test	0.080	19:17
Air Blank	0.000	19:18
Diagnostics Check	OK	19:18

Data Download Date/Time: 04/29/2016 11:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001291

INSTRUMENT REGISTERED TO: BRADFORD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SEAY, JOSHUA, W
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: SEAY, JOSHUA
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 02/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG425202
DRY GAS EXPIRY DATE: 09/09/2016
VIOLATION CODE: DUI
ARREST TIME: 01:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: MILLINGTON

DATE/TIME OF TEST: 03/03/2016 02:38:21

Diagnostics Check	OK	02:31
Air Blank	0.000	02:31
Control Test	0.080	02:31
Air Blank	0.000	02:32
Subject Sample # 1	0.132	02:33
Breath Volume	1.109	02:33
Air Blank	0.000	02:34
Air Blank	0.000	02:35
Subject Sample # 2	0.132	02:36
Breath Volume	2.992	02:36
Air Blank	0.000	02:37
Control Test	0.078	02:37
Air Blank	0.000	02:38
Diagnostics Check	OK	02:38

Data Download Date/Time: 03/18/2016 16:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001291

INSTRUMENT REGISTERED TO: BRADFORD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,TERRELL,G
BREATH TEST OPERATOR AGENCY: Bradford County SO
ARRESTING OFFICER: WILLIAMS,TERRELL
ARRESTING OFFICER AGENCY: Bradford County SO
LAST AGENCY INSPECTION: 02/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG425202
DRY GAS EXPIRY DATE: 09/09/2016
VIOLATION CODE: DUI
ARREST TIME: 17:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:55

SUBJECT'S LAST NAME: SZABO

DATE/TIME OF TEST: 03/03/2016 18:29:46

Diagnostics Check	OK	18:22
Air Blank	0.000	18:22
Control Test	0.078	18:22
Air Blank	0.000	18:23
Subject Sample # 1	0.091	18:25
Breath Volume	1.238	18:25
Air Blank	0.000	18:25
Air Blank	0.000	18:27
Subject Sample # 2	0.096	18:28
Breath Volume	1.628	18:28
Air Blank	0.000	18:28
Control Test	0.076	18:29
Air Blank	0.000	18:29
Diagnostics Check	OK	18:29

Data Download Date/Time: 03/18/2016 16:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001291

INSTRUMENT REGISTERED TO: BRADFORD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KONKEL,SCOTT,M
BREATH TEST OPERATOR AGENCY: Bradford County SO
ARRESTING OFFICER: GORDON,RUSSELL
ARRESTING OFFICER AGENCY: Bradford County SO
LAST AGENCY INSPECTION: 02/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG425202
DRY GAS EXPIRY DATE: 09/09/2016
VIOLATION CODE: DUI
ARREST TIME: 00:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: SHARP

DATE/TIME OF TEST: 03/05/2016 01:52:11

Diagnostics Check	OK	01:46
Air Blank	0.000	01:46
Control Test	0.078	01:46
Air Blank	0.000	01:47
Subject Sample # 1	0.177	01:47
Breath Volume	2.648	01:47
Air Blank	0.000	01:48
Air Blank	0.000	01:50
Subject Sample # 2	0.183	01:50
Breath Volume	3.910	01:50
Air Blank	0.000	01:51
Control Test	0.078	01:51
Air Blank	0.000	01:52
Diagnostics Check	OK	01:52

Data Download Date/Time: 03/18/2016 16:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001291

INSTRUMENT REGISTERED TO: BRADFORD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHILLER,JOSEPH,C
BREATH TEST OPERATOR AGENCY: Bradford County SO
ARRESTING OFFICER: MCREE,GARRETT
ARRESTING OFFICER AGENCY: Bradford County SO
LAST AGENCY INSPECTION: 02/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG425202
DRY GAS EXPIRY DATE: 09/09/2016
VIOLATION CODE: DUI
ARREST TIME: 10:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:45

SUBJECT'S LAST NAME: ROBBINS

DATE/TIME OF TEST: 03/09/2016 11:11:23

Diagnostics Check	OK	11:05
Air Blank	0.000	11:06
Control Test	0.079	11:06
Air Blank	0.000	11:06
Subject Sample # 1	0.000	11:07
Breath Volume	3.843	11:07
Air Blank	0.000	11:07
Air Blank	0.000	11:09
Subject Sample # 2	0.000	11:10
Breath Volume	3.156	11:10
Air Blank	0.000	11:10
Control Test	0.077	11:10
Air Blank	0.000	11:11
Diagnostics Check	OK	11:11

Data Download Date/Time: 03/18/2016 16:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001291

INSTRUMENT REGISTERED TO: BRADFORD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KONKEL,SCOTT,M
BREATH TEST OPERATOR AGENCY: Bradford County SO
ARRESTING OFFICER: CASSANI,DAN
ARRESTING OFFICER AGENCY: Bradford County SO
LAST AGENCY INSPECTION: 03/18/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG425202
DRY GAS EXPIRY DATE: 09/09/2016
VIOLATION CODE: DUI
ARREST TIME: 04:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:25

SUBJECT'S LAST NAME: STOUT

DATE/TIME OF TEST: 03/26/2016 04:51:49

Diagnostics Check	OK	04:45
Air Blank	0.000	04:45
Control Test	0.078	04:46
Air Blank	0.000	04:46
Subject Sample # 1	0.243	04:47
Breath Volume	1.964	04:47
Air Blank	0.000	04:47
Air Blank	0.000	04:49
Subject Sample # 2	0.235	04:50
Breath Volume	1.566	04:50
Air Blank	0.000	04:50
Control Test	0.077	04:51
Air Blank	0.000	04:51
Diagnostics Check	OK	04:51

Data Download Date/Time: 04/24/2016 09:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001291

INSTRUMENT REGISTERED TO: BRADFORD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KONKEL,SCOTT,M
BREATH TEST OPERATOR AGENCY: Bradford County SO
ARRESTING OFFICER: GORDON,RUSSELL
ARRESTING OFFICER AGENCY: Bradford County SO
LAST AGENCY INSPECTION: 03/18/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG425202
DRY GAS EXPIRY DATE: 09/09/2016
VIOLATION CODE: DUI
ARREST TIME: 20:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:20

SUBJECT'S LAST NAME: WEBSTER

DATE/TIME OF TEST: 03/30/2016 20:46:32

Diagnostics Check	OK	20:40
Air Blank	0.000	20:40
Control Test	0.079	20:40
Air Blank	0.000	20:41
Subject Sample # 1	0.267	20:41
Breath Volume	1.582	20:41
Air Blank	0.000	20:42
Air Blank	0.000	20:44
Subject Sample # 2	0.259	20:44
Breath Volume	1.703	20:44
Air Blank	0.000	20:45
Control Test	0.077	20:45
Air Blank	0.000	20:46
Diagnostics Check	OK	20:46

Data Download Date/Time: 04/24/2016 09:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPENCER, RAYMOND, L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: MADDEN, D
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 02/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 00:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:43

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 03/02/2016 04:07:27

Diagnostics Check	OK	04:03	
Air Blank	0.000	04:04	
Control Test	0.079	04:04	
Air Blank	0.000	04:05	
Subject Sample	# 1	04:06	Subject Test Refused
Breath Volume	0.000	04:06	
Air Blank	0.000	04:06	
Control Test	0.079	04:06	
Air Blank	0.000	04:07	
Diagnostics Check	OK	04:07	

Data Download Date/Time: 03/31/2016 12:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRUMBIR,EMERALD,G
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: GRUMBIR,EMERALD
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 02/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 03:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:24

SUBJECT'S LAST NAME: KLOEBER

DATE/TIME OF TEST: 03/05/2016 05:12:46

Diagnostics Check	OK	04:57	
Air Blank	0.000	04:58	
Control Test	0.080	04:58	
Air Blank	0.000	04:59	
Subject Sample	# 1	05:02	No Sample Provided
Breath Volume	0.000	05:02	
Air Blank	0.000	05:02	
Air Blank	0.000	05:04	
Subject Sample	# 2	0.212	05:05
Breath Volume	1.589	05:05	
Air Blank	0.000	05:06	
Air Blank	0.000	05:08	
Subject Sample	# 3	0.170	05:11
Breath Volume	0.648	05:11	Volume Not Met
Air Blank	0.000	05:11	
Control Test	0.078	05:12	
Air Blank	0.000	05:12	
Diagnostics Check	OK	05:12	

Data Download Date/Time: 03/31/2016 12:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRUMBIR,EMERALD,G
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: MADDEN,DWAYNE
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 02/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 03:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:34

SUBJECT'S LAST NAME: EASTWOOD

DATE/TIME OF TEST: 03/05/2016 05:39:08

Diagnostics Check	OK	05:28
Air Blank	0.000	05:29
Control Test	0.079	05:29
Air Blank	0.000	05:30
Subject Sample # 1	0.168	05:31
Breath Volume	4.488	05:31
Air Blank	0.000	05:32
Air Blank	0.000	05:34
Subject Sample # 2	0.142	05:34
Breath Volume	1.808	05:34
Air Blank	0.000	05:35
Air Blank	0.000	05:37
Subject Sample # 3	0.142	05:37
Breath Volume	1.429	05:37
Air Blank	0.000	05:38
Control Test	0.079	05:38
Air Blank	0.000	05:39
Diagnostics Check	OK	05:39

Data Download Date/Time: 03/31/2016 12:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CALABRO,ALFRED,J
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: CALABRO,ALFRED
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 02/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 23:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:21

SUBJECT'S LAST NAME: OLIVE

DATE/TIME OF TEST: 03/06/2016 00:47:09

Diagnostics Check	OK	00:43	
Air Blank	0.000	00:43	
Control Test	0.080	00:43	
Air Blank	0.000	00:44	
Subject Sample	# 1	00:45	Subject Test Refused
Breath Volume	0.000	00:45	
Air Blank	0.000	00:46	
Control Test	0.079	00:46	
Air Blank	0.000	00:47	
Diagnostics Check	OK	00:47	

Data Download Date/Time: 03/31/2016 12:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,COREY,M
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: JONES,COREY
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 02/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:57

SUBJECT'S LAST NAME: GIORGIO

DATE/TIME OF TEST: 03/06/2016 01:23:55

Diagnostics Check	OK	01:18
Air Blank	0.000	01:18
Control Test	0.080	01:18
Air Blank	0.000	01:19
Subject Sample # 1	0.000	01:19
Breath Volume	2.148	01:19
Air Blank	0.000	01:20
Air Blank	0.000	01:22
Subject Sample # 2	0.000	01:22
Breath Volume	2.285	01:22
Air Blank	0.000	01:23
Control Test	0.080	01:23
Air Blank	0.000	01:23
Diagnostics Check	OK	01:23

Data Download Date/Time: 03/31/2016 12:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AQUINO,OMAR,E
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: SHIVER,CHRISTOPHER
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 02/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 01:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:46

SUBJECT'S LAST NAME: VILLAIRE

DATE/TIME OF TEST: 03/06/2016 02:20:09

Diagnostics Check	OK	02:13
Air Blank	0.000	02:14
Control Test	0.079	02:14
Air Blank	0.000	02:15
Subject Sample # 1	0.254	02:15
Breath Volume	3.050	02:15
Air Blank	0.000	02:16
Air Blank	0.000	02:18
Subject Sample # 2	0.247	02:18
Breath Volume	2.480	02:18
Air Blank	0.000	02:19
Control Test	0.078	02:19
Air Blank	0.000	02:20
Diagnostics Check	OK	02:20

Data Download Date/Time: 03/31/2016 12:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRUMBIR,EMERALD,G
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: GRUMBIR,EMERALD
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 02/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 01:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: CLARK

DATE/TIME OF TEST: 03/06/2016 02:37:20

Diagnostics Check	OK	02:27
Air Blank	0.000	02:27
Control Test	0.079	02:28
Air Blank	0.000	02:28
Subject Sample # 1	0.138	02:30
Breath Volume	1.167	02:30
Air Blank	0.000	02:31
Air Blank	0.000	02:32
Subject Sample # 2	0.145	02:35
Breath Volume	1.417	02:35
Air Blank	0.000	02:36
Control Test	0.078	02:36
Air Blank	0.000	02:37
Diagnostics Check	OK	02:37

Data Download Date/Time: 03/31/2016 12:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,COREY,M
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: JONES,COREY
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 02/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 23:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:38

SUBJECT'S LAST NAME: GILLMAN

DATE/TIME OF TEST: 03/08/2016 01:01:58

Diagnostics Check	OK	00:59	
Air Blank	0.000	00:59	
Control Test	0.078	00:59	
Air Blank	0.000	01:00	
Subject Sample	# 1	01:00	Subject Test Refused
Breath Volume	0.000	01:00	
Air Blank	0.000	01:01	
Control Test	0.079	01:01	
Air Blank	0.000	01:01	
Diagnostics Check	OK	01:01	

Data Download Date/Time: 03/31/2016 12:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VITALE,MICHAEL,L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: VITALE,MICHAEL
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 02/26/2016

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:49

SUBJECT'S LAST NAME: WEBB

DATE/TIME OF TEST: 03/08/2016 04:16:06

Diagnostics Check	OK	04:10
Air Blank	0.000	04:10
Control Test	0.079	04:10
Air Blank	0.000	04:11
Subject Sample # 1	0.134	04:11
Breath Volume	1.417	04:11
Air Blank	0.000	04:12
Air Blank	0.000	04:14
Subject Sample # 2	0.143	04:14
Breath Volume	2.585	04:14
Air Blank	0.000	04:15
Control Test	0.079	04:15
Air Blank	0.000	04:16
Diagnostics Check	OK	04:16

Data Download Date/Time: 03/31/2016 12:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENDRIX,TIMOTHY,D
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: HENDRIX,TIMOTHY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 02/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 23:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: FRANKLIN

DATE/TIME OF TEST: 03/10/2016 00:16:21

Diagnostics Check	OK	00:09
Air Blank	0.000	00:09
Control Test	0.079	00:10
Air Blank	0.000	00:10
Subject Sample # 1	0.139	00:11
Breath Volume	1.621	00:11
Air Blank	0.000	00:12
Air Blank	0.000	00:14
Subject Sample # 2	0.139	00:14
Breath Volume	1.164	00:14
Air Blank	0.000	00:15
Control Test	0.079	00:15
Air Blank	0.000	00:16
Diagnostics Check	OK	00:16

Data Download Date/Time: 03/31/2016 12:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VITALE,MICHAEL,L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: VITALE,MICHAEL
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 02/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 01:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: CALDWELL

DATE/TIME OF TEST: 03/10/2016 02:41:45

Diagnostics Check	OK	02:35
Air Blank	0.000	02:36
Control Test	0.079	02:36
Air Blank	0.000	02:36
Subject Sample # 1	0.121	02:37
Breath Volume	2.882	02:37
Air Blank	0.000	02:37
Air Blank	0.000	02:39
Subject Sample # 2	0.114	02:40
Breath Volume	1.894	02:40
Air Blank	0.000	02:40
Control Test	0.078	02:41
Air Blank	0.000	02:41
Diagnostics Check	OK	02:41

Data Download Date/Time: 03/31/2016 12:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BEVIS,T,D
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BEVIS,TABITHA
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 02/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 13:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:40

SUBJECT'S LAST NAME: TEREKHIN

DATE/TIME OF TEST: 03/10/2016 15:20:06

Diagnostics Check	OK	15:13
Air Blank	0.000	15:14
Control Test	0.079	15:14
Air Blank	0.000	15:14
Subject Sample # 1	0.212	15:15
Breath Volume	1.921	15:15
Air Blank	0.000	15:16
Air Blank	0.000	15:17
Subject Sample # 2	0.192	15:18
Breath Volume	2.113	15:18
Air Blank	0.000	15:19
Control Test	0.076	15:19
Air Blank	0.000	15:20
Diagnostics Check	OK	15:20

Data Download Date/Time: 03/31/2016 12:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROWELL,DAVID,M
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: MADDEN,DWAYNE
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 02/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 23:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: ZIMA

DATE/TIME OF TEST: 03/11/2016 00:26:45

Diagnostics Check	OK	00:20
Air Blank	0.000	00:21
Control Test	0.079	00:21
Air Blank	0.000	00:21
Subject Sample # 1	0.057	00:22
Breath Volume	2.910	00:22
Air Blank	0.000	00:22
Air Blank	0.000	00:24
Subject Sample # 2	0.058	00:25
Breath Volume	3.992	00:25
Air Blank	0.000	00:25
Control Test	0.080	00:26
Air Blank	0.000	00:26
Diagnostics Check	OK	00:26

Data Download Date/Time: 03/31/2016 12:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHAU,LIEM,V
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: CHAU,LIEM
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 02/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 22:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:02

SUBJECT'S LAST NAME: MARQUEZ ALEMAN

DATE/TIME OF TEST: 03/12/2016 00:37:43

Diagnostics Check	OK	00:26	
Air Blank	0.000	00:26	
Control Test	0.080	00:27	
Air Blank	0.000	00:27	
Subject Sample	# 1	00:30	No Sample Provided
Breath Volume	0.000	00:30	
Air Blank	0.000	00:31	
Air Blank	0.000	00:33	
Subject Sample	# 2	00:36	No Sample Provided
Breath Volume	0.000	00:36	
Air Blank	0.000	00:36	
Control Test	0.080	00:37	
Air Blank	0.000	00:37	
Diagnostics Check	OK	00:37	

Data Download Date/Time: 03/31/2016 12:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CALABRO,ALFRED,J
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: CALABRO,ALFRED
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 02/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 23:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: BRUNER

DATE/TIME OF TEST: 03/12/2016 00:49:52

Diagnostics Check	OK	00:46	
Air Blank	0.000	00:46	
Control Test	0.079	00:47	
Air Blank	0.000	00:47	
Subject Sample	# 1	00:48	Subject Test Refused
Breath Volume	0.000	00:48	
Air Blank	0.000	00:48	
Control Test	0.079	00:49	
Air Blank	0.000	00:49	
Diagnostics Check	OK	00:49	

Data Download Date/Time: 03/31/2016 12:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENDRIX,TIMOTHY,D
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: HENDRIX,TIMOTHY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 02/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 00:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: MELICIA

DATE/TIME OF TEST: 03/12/2016 01:27:12

Diagnostics Check	OK	01:20
Air Blank	0.000	01:21
Control Test	0.079	01:21
Air Blank	0.000	01:21
Subject Sample # 1	0.222	01:22
Breath Volume	1.332	01:22
Air Blank	0.000	01:23
Air Blank	0.000	01:25
Subject Sample # 2	0.225	01:25
Breath Volume	1.437	01:25
Air Blank	0.000	01:26
Control Test	0.079	01:26
Air Blank	0.000	01:27
Diagnostics Check	OK	01:27

Data Download Date/Time: 03/31/2016 12:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VITALE,MICHAEL,L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: VITALE,MICHAEL
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 02/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:22

SUBJECT'S LAST NAME: HICKEY

DATE/TIME OF TEST: 03/12/2016 02:51:01

Diagnostics Check	OK	02:42	
Air Blank	0.000	02:42	
Control Test	0.080	02:43	
Air Blank	0.000	02:43	
Subject Sample	# 1 0.173	02:46	Volume Not Met
Breath Volume	0.687	02:46	
Air Blank	0.000	02:47	
Air Blank	0.000	02:49	
Subject Sample	# 2	02:49	Subject Test Refused
Breath Volume	0.000	02:49	
Air Blank	0.000	02:50	
Control Test	0.077	02:50	
Air Blank	0.000	02:50	
Diagnostics Check	OK	02:51	

Data Download Date/Time: 03/31/2016 12:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VITALE,MICHAEL,L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: VITALE,MICHAEL
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 02/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:22

SUBJECT'S LAST NAME: HICKEY

DATE/TIME OF TEST: 03/12/2016 02:59:37

Diagnostics Check	OK	02:52
Air Blank	0.000	02:52
Control Test	0.079	02:53
Air Blank	0.000	02:53
Subject Sample # 1	0.171	02:54
Breath Volume	1.531	02:54
Air Blank	0.000	02:54
Air Blank	0.000	02:56
Subject Sample # 2	0.171	02:58
Breath Volume	2.058	02:58
Air Blank	0.000	02:58
Control Test	0.079	02:59
Air Blank	0.000	02:59
Diagnostics Check	OK	02:59

Data Download Date/Time: 03/31/2016 12:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARVOTH,DANIEL,P
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: HARVOTH,DANIEL
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 02/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 23:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:47

SUBJECT'S LAST NAME: KOPKO

DATE/TIME OF TEST: 03/13/2016 01:22:42

Diagnostics Check		OK	01:09	
Air Blank		0.000	01:09	
Control Test		0.079	01:09	
Air Blank		0.000	01:10	
Subject Sample	# 1	0.190	01:13	Volume Not Met
Breath Volume		0.734	01:13	
Air Blank		0.000	01:14	
Air Blank		0.000	01:15	
Subject Sample	# 2	0.219	01:16	
Breath Volume		1.222	01:16	
Air Blank		0.000	01:17	
Air Blank		0.000	01:19	
Subject Sample	# 3	0.210	01:21	
Breath Volume		1.652	01:21	
Air Blank		0.000	01:21	
Control Test		0.078	01:22	
Air Blank		0.000	01:22	
Diagnostics Check		OK	01:22	

Data Download Date/Time: 03/31/2016 12:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUGHES,TIMOTHY,D
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: HUGHES,TIMOTHY
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 02/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 22:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: ;6360101856254091002

DATE/TIME OF TEST: 03/13/2016 23:41:51

Diagnostics Check	OK	23:38	
Air Blank	0.000	23:39	
Control Test	0.079	23:39	
Air Blank	0.000	23:40	
Subject Sample	# 1	23:40	Subject Test Refused
Breath Volume	0.000	23:40	
Air Blank	0.000	23:40	
Control Test	0.079	23:41	
Air Blank	0.000	23:41	
Diagnostics Check	OK	23:41	

Data Download Date/Time: 03/31/2016 12:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CALABRO,ALFRED,J
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: CALABRO,ALFRED
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 02/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 23:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: GARAY

DATE/TIME OF TEST: 03/15/2016 00:23:13

Diagnostics Check	OK	00:16
Air Blank	0.000	00:16
Control Test	0.080	00:17
Air Blank	0.000	00:17
Subject Sample # 1	0.127	00:18
Breath Volume	2.925	00:18
Air Blank	0.000	00:19
Air Blank	0.000	00:21
Subject Sample # 2	0.122	00:21
Breath Volume	2.335	00:21
Air Blank	0.000	00:22
Control Test	0.078	00:22
Air Blank	0.000	00:23
Diagnostics Check	OK	00:23

Data Download Date/Time: 03/31/2016 12:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROWELL,DAVID,M
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: ROWELL,DAVID
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 02/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: HOLSCLAW

DATE/TIME OF TEST: 03/15/2016 23:58:41

Diagnostics Check	OK	23:55	
Air Blank	0.000	23:56	
Control Test	0.079	23:56	
Air Blank	0.000	23:57	
Subject Sample	# 1	23:57	Subject Test Refused
Breath Volume	0.000	23:57	
Air Blank	0.000	23:57	
Control Test	0.079	23:58	
Air Blank	0.000	23:58	
Diagnostics Check	OK	23:58	

Data Download Date/Time: 03/31/2016 12:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPENCER, RAYMOND, L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: SPENCER, RAYMOND
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 02/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 01:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:01

SUBJECT'S LAST NAME: BURTTSCHELL

DATE/TIME OF TEST: 03/16/2016 02:33:44

Diagnostics Check		OK	02:22	
Air Blank		0.000	02:22	
Control Test		0.080	02:22	
Air Blank		0.000	02:23	
Subject Sample	# 1	0.249	02:26	Volume Not Met
Breath Volume		0.386	02:26	
Air Blank		0.000	02:27	
Air Blank		0.000	02:28	
Subject Sample	# 2	0.225	02:32	Volume Not Met
Breath Volume		0.187	02:32	
Air Blank		0.000	02:32	
Control Test		0.077	02:33	
Air Blank		0.000	02:33	
Diagnostics Check		OK	02:33	

Data Download Date/Time: 03/31/2016 12:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPENCER, RAYMOND, L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: SPENCER, RAYMOND
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 02/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 01:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:01

SUBJECT'S LAST NAME: BURTTSCHELL

DATE/TIME OF TEST: 03/16/2016 02:46:05

Diagnostics Check	OK	02:35	
Air Blank	0.000	02:36	
Control Test	0.077	02:36	
Air Blank	0.000	02:36	
Subject Sample	# 1 0.221	02:39	Volume Not Met
Breath Volume	1.050	02:39	
Air Blank	0.000	02:40	
Air Blank	0.000	02:42	
Subject Sample	# 2	02:44	Subject Test Refused
Breath Volume	0.031	02:44	
Air Blank	0.000	02:45	
Control Test	0.077	02:45	
Air Blank	0.000	02:46	
Diagnostics Check	OK	02:46	

Data Download Date/Time: 03/31/2016 12:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SWINDELL,SHAWN,T
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: GAGNON,JAMES
ARRESTING OFFICER AGENCY: Milton PD
LAST AGENCY INSPECTION: 02/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 08:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:18

SUBJECT'S LAST NAME: TASH

DATE/TIME OF TEST: 03/16/2016 09:44:25

Diagnostics Check	OK	09:38
Air Blank	0.000	09:38
Control Test	0.079	09:39
Air Blank	0.000	09:39
Subject Sample # 1	0.008	09:40
Breath Volume	1.757	09:40
Air Blank	0.000	09:40
Air Blank	0.000	09:42
Subject Sample # 2	0.007	09:43
Breath Volume	2.402	09:43
Air Blank	0.000	09:43
Control Test	0.080	09:43
Air Blank	0.000	09:44
Diagnostics Check	OK	09:44

Data Download Date/Time: 03/31/2016 12:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPENCER, RAYMOND, L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: SPENCER, RAYMOND
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 02/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 23:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:58

SUBJECT'S LAST NAME: ROBINSON

DATE/TIME OF TEST: 03/17/2016 00:27:11

Diagnostics Check		OK	00:18
Air Blank		0.000	00:19
Control Test		0.079	00:19
Air Blank		0.000	00:20
Subject Sample	# 1	0.200	00:21
Breath Volume		1.308	00:21
Air Blank		0.000	00:22
Air Blank		0.000	00:23
Subject Sample	# 2	0.199	00:25
Breath Volume		1.289	00:25
Air Blank		0.000	00:26
Control Test		0.079	00:26
Air Blank		0.000	00:27
Diagnostics Check		OK	00:27

Data Download Date/Time: 03/31/2016 12:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARTZOG,ROBERT,B
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: HARTZOG,ROBERT
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 02/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 21:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:21

SUBJECT'S LAST NAME: CALDWELL

DATE/TIME OF TEST: 03/17/2016 21:55:49

Diagnostics Check	OK	21:48
Air Blank	0.000	21:49
Control Test	0.079	21:49
Air Blank	0.000	21:50
Subject Sample # 1	0.159	21:51
Breath Volume	3.972	21:51
Air Blank	0.000	21:51
Air Blank	0.000	21:53
Subject Sample # 2	0.158	21:54
Breath Volume	3.066	21:54
Air Blank	0.000	21:54
Control Test	0.079	21:55
Air Blank	0.000	21:55
Diagnostics Check	OK	21:55

Data Download Date/Time: 03/31/2016 12:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACKSON,VICTOR,A
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: JACKSON,VICTOR
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 02/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 10:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: BOLES

DATE/TIME OF TEST: 03/18/2016 00:02:29

Diagnostics Check	OK	23:55
Air Blank	0.000	23:56
Control Test	0.078	23:56
Air Blank	0.000	23:56
Subject Sample # 2	0.215	23:57
Breath Volume	2.433	23:57
Air Blank	0.000	23:58
Air Blank	0.000	00:00
Subject Sample # 1	0.202	00:00
Breath Volume	2.492	00:00
Air Blank	0.000	00:01
Control Test	0.078	00:01
Air Blank	0.000	00:02
Diagnostics Check	OK	00:02

Data Download Date/Time: 03/31/2016 12:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BENNETT,EDGAR,A
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: BENNETT,EDGAR
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 02/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 01:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: TALARICO

DATE/TIME OF TEST: 03/18/2016 02:31:02

Diagnostics Check	OK	02:17		
Air Blank	0.000	02:18		
Control Test	0.079	02:18		
Air Blank	0.000	02:19		
Subject Sample	# 1	02:22	No Sample Provided	
Breath Volume	0.000	02:22		
Air Blank	0.000	02:22		
Air Blank	0.000	02:24		
Subject Sample	# 2	0.303	02:25	No .020 agreement
Breath Volume	1.218	02:25		
Air Blank	0.000	02:26		
Air Blank	0.000	02:28		
Subject Sample	# 3	0.279	02:29	No .020 agreement
Breath Volume	1.449	02:29		
Air Blank	0.000	02:30		
Control Test	0.077	02:30		
Air Blank	0.000	02:30		
Diagnostics Check	OK	02:31		

Data Download Date/Time: 03/31/2016 12:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AQUINO,OMAR,E
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: AQUINO,OMAR
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 02/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 01:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:07

SUBJECT'S LAST NAME: STEPHENS

DATE/TIME OF TEST: 03/18/2016 02:51:56

Diagnostics Check	OK	02:45
Air Blank	0.000	02:45
Control Test	0.077	02:46
Air Blank	0.000	02:46
Subject Sample # 1	0.227	02:47
Breath Volume	2.324	02:47
Air Blank	0.000	02:47
Air Blank	0.000	02:49
Subject Sample # 2	0.224	02:50
Breath Volume	2.441	02:50
Air Blank	0.000	02:51
Control Test	0.078	02:51
Air Blank	0.000	02:51
Diagnostics Check	OK	02:51

Data Download Date/Time: 03/31/2016 12:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AQUINO,OMAR,E
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: AQUINO,OMAR
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 02/26/2016

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 12:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:28

SUBJECT'S LAST NAME: JERNIGAN

DATE/TIME OF TEST: 03/19/2016 01:59:14

Diagnostics Check	OK	01:52
Air Blank	0.000	01:52
Control Test	0.079	01:52
Air Blank	0.000	01:53
Subject Sample # 1	0.229	01:54
Breath Volume	1.628	01:54
Air Blank	0.000	01:54
Air Blank	0.000	01:56
Subject Sample # 2	0.228	01:57
Breath Volume	1.828	01:57
Air Blank	0.000	01:58
Control Test	0.079	01:58
Air Blank	0.000	01:59
Diagnostics Check	OK	01:59

Data Download Date/Time: 03/31/2016 12:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COOK,CALEB,Z
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: COOK,CALEB
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 02/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 15:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:54

SUBJECT'S LAST NAME: COLLINS

DATE/TIME OF TEST: 03/19/2016 17:24:04

Diagnostics Check	OK	17:17
Air Blank	0.000	17:17
Control Test	0.078	17:18
Air Blank	0.000	17:18
Subject Sample # 1	0.187	17:19
Breath Volume	2.832	17:19
Air Blank	0.000	17:20
Air Blank	0.000	17:21
Subject Sample # 2	0.180	17:22
Breath Volume	2.023	17:22
Air Blank	0.000	17:23
Control Test	0.078	17:23
Air Blank	0.000	17:23
Diagnostics Check	OK	17:24

Data Download Date/Time: 03/31/2016 12:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARLSON,JOSHUA,L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: CARLSON,JOSHUA
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 02/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 16:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:50

SUBJECT'S LAST NAME: ROBEY

DATE/TIME OF TEST: 03/19/2016 17:33:42

Diagnostics Check	OK	17:26
Air Blank	0.000	17:27
Control Test	0.078	17:27
Air Blank	0.000	17:28
Subject Sample # 1	0.142	17:28
Breath Volume	2.707	17:28
Air Blank	0.000	17:29
Air Blank	0.000	17:31
Subject Sample # 2	0.140	17:32
Breath Volume	1.238	17:32
Air Blank	0.000	17:32
Control Test	0.078	17:33
Air Blank	0.000	17:33
Diagnostics Check	OK	17:33

Data Download Date/Time: 03/31/2016 12:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMPSON,ANDREW,J
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: SIMPSON,ANDREW
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 02/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 08:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:23

SUBJECT'S LAST NAME: GARRETT

DATE/TIME OF TEST: 03/19/2016 21:46:22

Diagnostics Check	OK	21:43	
Air Blank	0.000	21:43	
Control Test	0.079	21:44	
Air Blank	0.000	21:44	
Subject Sample	# 1	21:45	Subject Test Refused
Breath Volume	0.000	21:45	
Air Blank	0.000	21:45	
Control Test	0.079	21:45	
Air Blank	0.000	21:46	
Diagnostics Check	OK	21:46	

Data Download Date/Time: 03/31/2016 12:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOBBS,JUSTIN,E
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: HOBBS,JUSTIN
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 02/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 21:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:25

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 03/19/2016 23:15:52

Diagnostics Check	OK	23:08
Air Blank	0.000	23:08
Control Test	0.080	23:08
Air Blank	0.000	23:09
Subject Sample # 1	0.137	23:10
Breath Volume	3.554	23:10
Air Blank	0.000	23:10
Air Blank	0.000	23:12
Subject Sample # 2	0.137	23:14
Breath Volume	3.718	23:14
Air Blank	0.000	23:14
Control Test	0.080	23:15
Air Blank	0.000	23:15
Diagnostics Check	OK	23:15

Data Download Date/Time: 03/31/2016 12:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BATSON,JAMES,H
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: BATSON,JAMES
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 02/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 00:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:08

SUBJECT'S LAST NAME: DUPONT

DATE/TIME OF TEST: 03/20/2016 01:34:53

Diagnostics Check	OK	01:28
Air Blank	0.000	01:28
Control Test	0.080	01:29
Air Blank	0.000	01:29
Subject Sample # 1	0.199	01:30
Breath Volume	3.003	01:30
Air Blank	0.000	01:30
Air Blank	0.000	01:32
Subject Sample # 2	0.196	01:33
Breath Volume	2.847	01:33
Air Blank	0.000	01:33
Control Test	0.079	01:34
Air Blank	0.000	01:34
Diagnostics Check	OK	01:34

Data Download Date/Time: 03/31/2016 12:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COUNTS,RYAN,L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: COUNTS,RYAN
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 02/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:18

SUBJECT'S LAST NAME: JAMET

DATE/TIME OF TEST: 03/20/2016 03:51:50

Diagnostics Check	OK	03:43	
Air Blank	0.000	03:43	
Control Test	0.080	03:44	
Air Blank	0.000	03:44	
Subject Sample	# 1	03:47	No Sample Provided
Breath Volume	0.000	03:47	
Air Blank	0.000	03:48	
Air Blank	0.000	03:50	
Subject Sample	# 2	03:50	Subject Test Refused
Breath Volume	0.000	03:50	
Air Blank	0.000	03:50	
Control Test	0.079	03:51	
Air Blank	0.000	03:51	
Diagnostics Check	OK	03:51	

Data Download Date/Time: 03/31/2016 12:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VITALE,MICHAEL,L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: VITALE,MICHAEL
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 02/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: BARNER

DATE/TIME OF TEST: 03/24/2016 01:17:03

Diagnostics Check	OK	01:10
Air Blank	0.000	01:11
Control Test	0.080	01:11
Air Blank	0.000	01:12
Subject Sample # 1	0.125	01:12
Breath Volume	2.886	01:12
Air Blank	0.000	01:13
Air Blank	0.000	01:15
Subject Sample # 2	0.126	01:15
Breath Volume	2.828	01:15
Air Blank	0.000	01:16
Control Test	0.080	01:16
Air Blank	0.000	01:16
Diagnostics Check	OK	01:17

Data Download Date/Time: 03/31/2016 12:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROWELL,DAVID,M
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: ROWELL,DAVID
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 02/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 01:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:23

SUBJECT'S LAST NAME: HUFFMAN

DATE/TIME OF TEST: 03/24/2016 01:50:31

Diagnostics Check	OK	01:43
Air Blank	0.000	01:44
Control Test	0.080	01:44
Air Blank	0.000	01:45
Subject Sample # 1	0.188	01:46
Breath Volume	3.046	01:46
Air Blank	0.000	01:46
Air Blank	0.000	01:48
Subject Sample # 2	0.189	01:48
Breath Volume	2.984	01:48
Air Blank	0.000	01:49
Control Test	0.079	01:49
Air Blank	0.000	01:50
Diagnostics Check	OK	01:50

Data Download Date/Time: 03/31/2016 12:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,JOHN,F
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: BERRIAN,ALTON
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 02/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 02:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: ROSARIO GERENA

DATE/TIME OF TEST: 03/25/2016 04:12:46

Diagnostics Check	OK	04:01	
Air Blank	0.000	04:01	
Control Test	0.079	04:01	
Air Blank	0.000	04:02	
Subject Sample	# 1 0.192	04:05	Volume Not Met
Breath Volume	0.671	04:05	
Air Blank	0.000	04:06	
Air Blank	0.000	04:08	
Subject Sample	# 2 0.146	04:11	Volume Not Met
Breath Volume	0.687	04:11	
Air Blank	0.000	04:11	
Control Test	0.077	04:12	
Air Blank	0.000	04:12	
Diagnostics Check	OK	04:12	

Data Download Date/Time: 03/31/2016 12:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COUNTS,RYAN,L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: COUNTS,RYAN
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 02/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 18:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:30

SUBJECT'S LAST NAME: DOWELL

DATE/TIME OF TEST: 03/25/2016 19:59:22

Diagnostics Check	OK	19:50	
Air Blank	0.000	19:51	
Control Test	0.079	19:51	
Air Blank	0.000	19:52	
Subject Sample	# 1	19:55	No Sample Provided
Breath Volume	0.000	19:55	
Air Blank	0.000	19:55	
Air Blank	0.000	19:57	
Subject Sample	# 2	19:58	Subject Test Refused
Breath Volume	0.000	19:58	
Air Blank	0.000	19:58	
Control Test	0.079	19:58	
Air Blank	0.000	19:59	
Diagnostics Check	OK	19:59	

Data Download Date/Time: 03/31/2016 12:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VITALE,MICHAEL,L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: VITALE,MICHAEL
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 02/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 02:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:58

SUBJECT'S LAST NAME: PATTERSON

DATE/TIME OF TEST: 03/27/2016 03:26:00

Diagnostics Check	OK	03:19
Air Blank	0.000	03:19
Control Test	0.080	03:20
Air Blank	0.000	03:20
Subject Sample # 1	0.144	03:21
Breath Volume	1.597	03:21
Air Blank	0.000	03:22
Air Blank	0.000	03:24
Subject Sample # 2	0.154	03:24
Breath Volume	1.277	03:24
Air Blank	0.000	03:25
Control Test	0.079	03:25
Air Blank	0.000	03:25
Diagnostics Check	OK	03:26

Data Download Date/Time: 03/31/2016 12:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPENCER, RAYMOND, L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: SPENCER, RAYMOND
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 02/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 04:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: WALLACE

DATE/TIME OF TEST: 03/27/2016 04:54:19

Diagnostics Check	OK	04:50	
Air Blank	0.000	04:51	
Control Test	0.080	04:51	
Air Blank	0.000	04:52	
Subject Sample	# 1	04:52	Subject Test Refused
Breath Volume	0.000	04:52	
Air Blank	0.000	04:53	
Control Test	0.079	04:53	
Air Blank	0.000	04:54	
Diagnostics Check	OK	04:54	

Data Download Date/Time: 03/31/2016 12:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRUMBIR,EMERALD,G
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: GRUMBIR,EMERALD
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 02/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 00:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: MORRISON

DATE/TIME OF TEST: 03/30/2016 02:03:49

Diagnostics Check	OK	01:57
Air Blank	0.000	01:57
Control Test	0.080	01:57
Air Blank	0.000	01:58
Subject Sample # 1	0.163	01:59
Breath Volume	3.179	01:59
Air Blank	0.000	01:59
Air Blank	0.000	02:01
Subject Sample # 2	0.157	02:02
Breath Volume	2.984	02:02
Air Blank	0.000	02:02
Control Test	0.078	02:03
Air Blank	0.000	02:03
Diagnostics Check	OK	02:03

Data Download Date/Time: 03/31/2016 12:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MANN,CHRISTOPHER,M
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: MANN,CHRISTOPHER
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 02/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 02:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: MCARTHUR

DATE/TIME OF TEST: 03/31/2016 03:10:48

Diagnostics Check	OK	03:07	
Air Blank	0.000	03:08	
Control Test	0.079	03:08	
Air Blank	0.000	03:08	
Subject Sample	# 1	03:09	Subject Test Refused
Breath Volume	0.000	03:09	
Air Blank	0.000	03:09	
Control Test	0.079	03:10	
Air Blank	0.000	03:10	
Diagnostics Check	OK	03:10	

Data Download Date/Time: 03/31/2016 12:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VITALE,MICHAEL,L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: VITALE,MICHAEL
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 02/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 04:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:56

SUBJECT'S LAST NAME: ROLLO

DATE/TIME OF TEST: 03/31/2016 05:20:01

Diagnostics Check	OK	05:16	
Air Blank	0.000	05:16	
Control Test	0.080	05:17	
Air Blank	0.000	05:17	
Subject Sample	# 1	05:18	Subject Test Refused
Breath Volume	0.000	05:18	
Air Blank	0.000	05:19	
Control Test	0.080	05:19	
Air Blank	0.000	05:19	
Diagnostics Check	OK	05:20	

Data Download Date/Time: 03/31/2016 12:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CALABRO,ALFRED,J
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: CALABRO,ALFRED
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 02/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 04:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:40

SUBJECT'S LAST NAME: SMILEY

DATE/TIME OF TEST: 03/31/2016 06:03:43

Diagnostics Check	OK	06:00	
Air Blank	0.000	06:01	
Control Test	0.079	06:01	
Air Blank	0.000	06:02	
Subject Sample	# 1	06:02	Subject Test Refused
Breath Volume	0.000	06:02	
Air Blank	0.000	06:02	
Control Test	0.079	06:03	
Air Blank	0.000	06:03	
Diagnostics Check	OK	06:03	

Data Download Date/Time: 03/31/2016 12:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001305

INSTRUMENT REGISTERED TO: MARIANNA PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: RILEY,DAVID,L

BREATH TEST OPERATOR AGENCY: Marianna PD

ARRESTING OFFICER: KIMMICH,MARTY

ARRESTING OFFICER AGENCY: Marianna PD

LAST AGENCY INSPECTION: 02/19/2016

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1650824

DRY GAS EXPIRY DATE: 03/13/2017

VIOLATION CODE: DUI

ARREST TIME: 00:48

GENDER: Male

OBSERVATION PERIOD BEGAN: 00:58

SUBJECT'S LAST NAME: GAINEY

DATE/TIME OF TEST: 03/16/2016 01:25:15

Diagnostics Check	OK	01:22	
Air Blank	0.000	01:22	
Control Test	0.078	01:23	
Air Blank	0.000	01:23	
Subject Sample	# 1	01:23	Subject Test Refused
Breath Volume	0.000	01:23	
Air Blank	0.000	01:24	
Control Test	0.078	01:24	
Air Blank	0.000	01:25	
Diagnostics Check	OK	01:25	

Data Download Date/Time: 03/23/2016 20:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001307

INSTRUMENT REGISTERED TO: MADISON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAKER,MARK,J
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: BAKER,MARK
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 02/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20314080A4
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 13:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:10

SUBJECT'S LAST NAME: BROYLES

DATE/TIME OF TEST: 03/05/2016 15:45:32

Diagnostics Check	OK	15:37
Air Blank	0.000	15:37
Control Test	0.079	15:38
Air Blank	0.000	15:38
Subject Sample # 1	0.104	15:40
Breath Volume	1.429	15:40
Air Blank	0.000	15:41
Air Blank	0.000	15:43
Subject Sample # 2	0.101	15:44
Breath Volume	1.742	15:44
Air Blank	0.000	15:44
Control Test	0.078	15:44
Air Blank	0.000	15:45
Diagnostics Check	OK	15:45

Data Download Date/Time: 03/08/2016 10:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAMMONS,CHRISTOPHER,R
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: WILSON,LAMAR
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 02/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG507503
DRY GAS EXPIRY DATE: 03/16/2017
VIOLATION CODE: DUI
ARREST TIME: 04:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:35

SUBJECT'S LAST NAME: BALLARD

DATE/TIME OF TEST: 03/01/2016 05:12:50

Diagnostics Check	OK	05:06
Air Blank	0.000	05:07
Control Test	0.081	05:07
Air Blank	0.000	05:08
Subject Sample # 1	0.000	05:08
Breath Volume	3.386	05:08
Air Blank	0.000	05:08
Air Blank	0.000	05:10
Subject Sample # 2	0.000	05:11
Breath Volume	3.402	05:11
Air Blank	0.000	05:11
Control Test	0.081	05:12
Air Blank	0.000	05:12
Diagnostics Check	OK	05:12

Data Download Date/Time: 03/28/2016 11:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURNETTE,AMY,M
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: SCHELL,GARY
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 02/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG507503
DRY GAS EXPIRY DATE: 03/16/2017
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:50

SUBJECT'S LAST NAME: OBERMILLER

DATE/TIME OF TEST: 03/06/2016 05:21:32

Diagnostics Check	OK	05:13
Air Blank	0.000	05:14
Control Test	0.079	05:14
Air Blank	0.000	05:15
Subject Sample # 1	0.225	05:16
Breath Volume	2.335	05:16
Air Blank	0.000	05:17
Air Blank	0.000	05:19
Subject Sample # 2	0.208	05:19
Breath Volume	1.187	05:19
Air Blank	0.000	05:20
Control Test	0.078	05:20
Air Blank	0.000	05:21
Diagnostics Check	OK	05:21

Data Download Date/Time: 03/28/2016 11:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAMMONS,CHRISTOPHER,R
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: CIANELL,V
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 02/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG507503
DRY GAS EXPIRY DATE: 03/16/2017
VIOLATION CODE: DUI
ARREST TIME: 23:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: MERCER

DATE/TIME OF TEST: 03/07/2016 00:28:27

Diagnostics Check		OK	00:20
Air Blank		0.000	00:21
Control Test		0.080	00:21
Air Blank		0.000	00:22
Subject Sample	# 1	0.210	00:23
Breath Volume		1.894	00:23
Air Blank		0.000	00:24
Air Blank		0.000	00:25
Subject Sample	# 2	0.199	00:26
Breath Volume		2.191	00:26
Air Blank		0.000	00:27
Control Test		0.079	00:27
Air Blank		0.000	00:28
Diagnostics Check		OK	00:28

Data Download Date/Time: 03/28/2016 11:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAMMONS,CHRISTOPHER,R
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: PRICE,C
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 02/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG507503
DRY GAS EXPIRY DATE: 03/16/2017
VIOLATION CODE: DUI
ARREST TIME: 23:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:48

SUBJECT'S LAST NAME: TRENT

DATE/TIME OF TEST: 03/07/2016 00:55:00

Diagnostics Check	OK	00:52	
Air Blank	0.000	00:52	
Control Test	0.080	00:52	
Air Blank	0.000	00:53	
Subject Sample	# 1	00:53	Subject Test Refused
Breath Volume	0.000	00:53	
Air Blank	0.000	00:54	
Control Test	0.080	00:54	
Air Blank	0.000	00:54	
Diagnostics Check	OK	00:55	

Data Download Date/Time: 03/28/2016 11:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAMMONS,CHRISTOPHER,R
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: STRIKLAND,OFFICER
ARRESTING OFFICER AGENCY: Parker PD
LAST AGENCY INSPECTION: 02/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG507503
DRY GAS EXPIRY DATE: 03/16/2017
VIOLATION CODE: DUI
ARREST TIME: 00:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: LAMBERT

DATE/TIME OF TEST: 03/11/2016 01:12:25

Diagnostics Check	OK	01:05
Air Blank	0.000	01:06
Control Test	0.080	01:06
Air Blank	0.000	01:07
Subject Sample # 1	0.120	01:07
Breath Volume	2.718	01:07
Air Blank	0.000	01:08
Air Blank	0.000	01:10
Subject Sample # 2	0.117	01:10
Breath Volume	1.437	01:10
Air Blank	0.000	01:11
Control Test	0.080	01:11
Air Blank	0.000	01:12
Diagnostics Check	OK	01:12

Data Download Date/Time: 03/28/2016 11:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROSIUS,STANLEY,W
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: WILLOUGHBY,R
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 02/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG507503
DRY GAS EXPIRY DATE: 03/16/2017
VIOLATION CODE: DUI
ARREST TIME: 22:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:17

SUBJECT'S LAST NAME: RICHARD

DATE/TIME OF TEST: 03/11/2016 23:49:52

Diagnostics Check	OK	23:42
Air Blank	0.000	23:43
Control Test	0.080	23:43
Air Blank	0.000	23:44
Subject Sample # 1	0.233	23:44
Breath Volume	2.484	23:44
Air Blank	0.000	23:45
Air Blank	0.000	23:47
Subject Sample # 2	0.230	23:48
Breath Volume	2.128	23:48
Air Blank	0.000	23:48
Control Test	0.080	23:49
Air Blank	0.000	23:49
Diagnostics Check	OK	23:49

Data Download Date/Time: 03/28/2016 11:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUBEL,NEIL,F
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: ALDERMAN,JOSEPH
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 02/26/2016

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: AG507503
DRY GAS EXPIRY DATE: 03/16/2017
VIOLATION CODE: DUI
ARREST TIME: 12:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 12:51

SUBJECT'S LAST NAME: LAWRENCE

DATE/TIME OF TEST: 03/14/2016 13:26:30

Diagnostics Check		OK	13:12	
Air Blank		0.000	13:12	
Control Test		0.079	13:12	
Air Blank		0.000	13:13	
Subject Sample	# 1	0.347	13:16	Volume Not Met
Breath Volume		0.406	13:16	
Air Blank		0.000	13:17	
Air Blank		0.000	13:19	
Subject Sample	# 2	0.354	13:21	No .020 agreement
Breath Volume		1.109	13:21	
Air Blank		0.000	13:22	
Air Blank		0.000	13:23	
Subject Sample	# 3	0.395	13:24	No .020 agreement
Breath Volume		1.488	13:24	
Air Blank		0.000	13:25	
Control Test		0.077	13:25	
Air Blank		0.000	13:26	
Diagnostics Check		OK	13:26	

Data Download Date/Time: 03/28/2016 11:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUBEL,NEIL,F
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: ALDERMAN,JOSEPH
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 02/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG507503
DRY GAS EXPIRY DATE: 03/16/2017
VIOLATION CODE: DUI
ARREST TIME: 12:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 12:51

SUBJECT'S LAST NAME: LAWRENCE

DATE/TIME OF TEST: 03/14/2016 13:42:10

Diagnostics Check		OK	13:30	
Air Blank		0.000	13:30	
Control Test		0.078	13:30	
Air Blank		0.000	13:31	
Subject Sample	# 1	0.332	13:34	Volume Not Met
Breath Volume		0.878	13:34	
Air Blank		0.000	13:35	
Air Blank		0.000	13:36	
Subject Sample	# 2	0.411	13:37	
Breath Volume		1.839	13:37	
Air Blank		0.000	13:38	
Air Blank		0.000	13:39	
Subject Sample	# 3	0.411	13:40	
Breath Volume		1.375	13:40	
Air Blank		0.000	13:41	
Control Test		0.078	13:41	
Air Blank		0.000	13:42	
Diagnostics Check		OK	13:42	

Data Download Date/Time: 03/28/2016 11:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAMMONS,CHRISTOPER,R
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: SCHELL,G
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 02/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG507503
DRY GAS EXPIRY DATE: 03/16/2017
VIOLATION CODE: DUI
ARREST TIME: 01:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:53

SUBJECT'S LAST NAME: HOWARD

DATE/TIME OF TEST: 03/19/2016 02:20:07

Diagnostics Check	OK	02:13
Air Blank	0.000	02:14
Control Test	0.079	02:14
Air Blank	0.000	02:15
Subject Sample # 1	0.090	02:15
Breath Volume	2.082	02:15
Air Blank	0.000	02:16
Air Blank	0.000	02:18
Subject Sample # 2	0.087	02:18
Breath Volume	1.875	02:18
Air Blank	0.000	02:19
Control Test	0.080	02:19
Air Blank	0.000	02:20
Diagnostics Check	OK	02:20

Data Download Date/Time: 03/28/2016 11:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLYN,WILLIAM,P
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: ALLYN,WILLIAM
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 02/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG507503
DRY GAS EXPIRY DATE: 03/16/2017
VIOLATION CODE: DUI
ARREST TIME: 00:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: MORRIS

DATE/TIME OF TEST: 03/22/2016 01:58:47

Diagnostics Check		OK	01:52
Air Blank		0.000	01:52
Control Test		0.081	01:53
Air Blank		0.000	01:53
Subject Sample	# 1	0.204	01:54
Breath Volume		4.156	01:54
Air Blank		0.000	01:54
Air Blank		0.000	01:56
Subject Sample	# 2	0.199	01:57
Breath Volume		4.503	01:57
Air Blank		0.000	01:57
Control Test		0.080	01:58
Air Blank		0.000	01:58
Diagnostics Check		OK	01:58

Data Download Date/Time: 03/28/2016 11:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAMMONS,CHRISTOPHER,R
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: CIESLIK,J
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 02/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG507503
DRY GAS EXPIRY DATE: 03/16/2017
VIOLATION CODE: DUI
ARREST TIME: 19:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:30

SUBJECT'S LAST NAME: LINK

DATE/TIME OF TEST: 03/23/2016 19:53:37

Diagnostics Check	OK	19:50	
Air Blank	0.000	19:51	
Control Test	0.080	19:51	
Air Blank	0.000	19:52	
Subject Sample	# 1	19:52	Subject Test Refused
Breath Volume	0.000	19:52	
Air Blank	0.000	19:52	
Control Test	0.080	19:53	
Air Blank	0.000	19:53	
Diagnostics Check	OK	19:53	

Data Download Date/Time: 03/28/2016 11:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHAW,MALIK,S
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: JOHNSON,OFC
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 02/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG507503
DRY GAS EXPIRY DATE: 03/16/2017
VIOLATION CODE: DUI
ARREST TIME: 13:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:18

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 03/25/2016 14:17:30

Diagnostics Check	OK	14:13	
Air Blank	0.000	14:14	
Control Test	0.080	14:14	
Air Blank	0.000	14:15	
Subject Sample	# 1	14:16	Subject Test Refused
Breath Volume	0.000	14:16	
Air Blank	0.000	14:16	
Control Test	0.080	14:16	
Air Blank	0.000	14:17	
Diagnostics Check	OK	14:17	

Data Download Date/Time: 03/28/2016 11:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOTURCO,DEPUTY,
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: CHERI,DEPUTY M
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 02/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG507503
DRY GAS EXPIRY DATE: 03/16/2017
VIOLATION CODE: DUI
ARREST TIME: 00:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:26

SUBJECT'S LAST NAME: CARBAJAL

DATE/TIME OF TEST: 03/27/2016 01:53:37

Diagnostics Check		OK	01:47
Air Blank		0.000	01:47
Control Test		0.080	01:47
Air Blank		0.000	01:48
Subject Sample	# 1	0.251	01:48
Breath Volume		1.929	01:48
Air Blank		0.000	01:49
Air Blank		0.000	01:51
Subject Sample	# 2	0.249	01:51
Breath Volume		3.210	01:51
Air Blank		0.000	01:52
Control Test		0.078	01:53
Air Blank		0.000	01:53
Diagnostics Check		OK	01:53

Data Download Date/Time: 03/28/2016 11:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TARRANTS,JEFFERY,K
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: PIERCE,DOUG
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 02/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG507503
DRY GAS EXPIRY DATE: 03/16/2017
VIOLATION CODE: DUI
ARREST TIME: 08:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:50

SUBJECT'S LAST NAME: PARSONS

DATE/TIME OF TEST: 03/14/2016 09:17:25

Diagnostics Check	OK	09:12	
Air Blank	0.000	09:12	
Control Test	0.079	09:12	
Air Blank	0.000	09:13	
Subject Sample	# 1 0.057	09:13	
Breath Volume	1.949	09:13	
Air Blank	0.000	09:14	
Air Blank	0.000	09:16	
Subject Sample	# 2	09:16	Sequence Aborted
Breath Volume	0.000	09:16	
Air Blank	0.000	09:17	

Data Download Date/Time: 03/28/2016 11:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TARRANTS,JEFFERY,K
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: PIERCE,DOUG
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 02/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG507503
DRY GAS EXPIRY DATE: 03/16/2017
VIOLATION CODE: DUI
ARREST TIME: 08:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:50

SUBJECT'S LAST NAME: PARSONS

DATE/TIME OF TEST: 03/14/2016 09:28:46

Diagnostics Check	OK	09:21
Air Blank	0.000	09:22
Control Test	0.077	09:22
Air Blank	0.000	09:22
Subject Sample # 1	0.052	09:23
Breath Volume	1.164	09:23
Air Blank	0.000	09:24
Air Blank	0.000	09:26
Subject Sample # 2	0.056	09:27
Breath Volume	2.597	09:27
Air Blank	0.000	09:27
Control Test	0.078	09:28
Air Blank	0.000	09:28
Diagnostics Check	OK	09:28

Data Download Date/Time: 03/28/2016 11:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001310

INSTRUMENT REGISTERED TO: NICEVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LITTLE, COLLETTE, A
BREATH TEST OPERATOR AGENCY: Niceville PD
ARRESTING OFFICER: SULLIVAN, DANNY
ARRESTING OFFICER AGENCY: Niceville PD
LAST AGENCY INSPECTION: 02/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12014080A1
DRY GAS EXPIRY DATE: 06/01/2016
VIOLATION CODE: DUI
ARREST TIME: 08:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:46

SUBJECT'S LAST NAME: BOYKIN

DATE/TIME OF TEST: 03/09/2016 09:17:37

Diagnostics Check	OK	09:10	
Air Blank	0.000	09:10	
Control Test	0.075	09:10	
Air Blank	0.000	09:11	
Subject Sample # 1	0.304	09:12	
Breath Volume	1.113	09:12	
Air Blank	0.000	09:13	
Air Blank	0.000	09:15	
Subject Sample # 2	0.292	09:16	
Breath Volume	1.429	09:16	
Air Blank	0.000	09:16	
Control Test	0.073	09:17	Control Outside Tolerance
Air Blank	0.000	09:17	

Data Download Date/Time: 03/31/2016 04:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001310

INSTRUMENT REGISTERED TO: NICEVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LITTLE, COLLETTE, A
BREATH TEST OPERATOR AGENCY: Niceville PD
ARRESTING OFFICER: SULLIVAN, DANNY
ARRESTING OFFICER AGENCY: Niceville PD
LAST AGENCY INSPECTION: 02/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29614080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 08:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:46

SUBJECT'S LAST NAME: BOYKIN

DATE/TIME OF TEST: 03/09/2016 10:02:10

Diagnostics Check		OK	09:55
Air Blank		0.000	09:56
Control Test		0.079	09:56
Air Blank		0.000	09:57
Subject Sample	# 1	0.286	09:57
Breath Volume		1.945	09:57
Air Blank		0.000	09:58
Air Blank		0.000	10:00
Subject Sample	# 2	0.272	10:00
Breath Volume		1.777	10:00
Air Blank		0.000	10:01
Control Test		0.078	10:01
Air Blank		0.000	10:02
Diagnostics Check		OK	10:02

Data Download Date/Time: 03/31/2016 04:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEWIS,TABITHA,J
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: MOUNT,DEBORAH
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 03/01/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 21:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:38

SUBJECT'S LAST NAME: LYNCH

DATE/TIME OF TEST: 03/02/2016 23:04:36

Diagnostics Check	OK	23:00	
Air Blank	0.000	23:01	
Control Test	0.083	23:01	
Air Blank	0.000	23:02	
Subject Sample	# 1	23:03	Subject Test Refused
Breath Volume	0.000	23:03	
Air Blank	0.000	23:03	
Control Test	0.082	23:03	
Air Blank	0.000	23:04	
Diagnostics Check	OK	23:04	

Data Download Date/Time: 04/27/2016 11:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RENNA,DAVID,P
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: N/A,N/A
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 03/01/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: PROB/PAROL
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 03/08/2016 00:34:28

Diagnostics Check	OK	00:28
Air Blank	0.000	00:28
Control Test	0.083	00:29
Air Blank	0.000	00:29
Subject Sample # 1	0.000	00:30
Breath Volume	4.523	00:30
Air Blank	0.000	00:30
Air Blank	0.000	00:32
Subject Sample # 2	0.000	00:33
Breath Volume	3.605	00:33
Air Blank	0.000	00:33
Control Test	0.081	00:33
Air Blank	0.000	00:34
Diagnostics Check	OK	00:34

Data Download Date/Time: 04/27/2016 12:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AIMANOVICH, MATTHEW, J
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: N/A, N/A
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 03/01/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: PROB/PAROL
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:56

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 03/10/2016 18:23:37

Diagnostics Check	OK	18:17
Air Blank	0.000	18:17
Control Test	0.082	18:18
Air Blank	0.000	18:18
Subject Sample # 1	0.000	18:19
Breath Volume	3.722	18:19
Air Blank	0.000	18:19
Air Blank	0.000	18:21
Subject Sample # 2	0.000	18:22
Breath Volume	4.250	18:22
Air Blank	0.000	18:22
Control Test	0.082	18:23
Air Blank	0.000	18:23
Diagnostics Check	OK	18:23

Data Download Date/Time: 04/27/2016 12:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBINSON,CHRISTOPHER,A
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: ROBINSON,CHRISTOPHER
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 03/01/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 01:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: PARKER

DATE/TIME OF TEST: 03/12/2016 02:46:59

Diagnostics Check	OK	02:35	
Air Blank	0.000	02:36	
Control Test	0.082	02:36	
Air Blank	0.000	02:36	
Subject Sample	# 1	02:40	No Sample Provided
Breath Volume	0.000	02:40	
Air Blank	0.000	02:40	
Air Blank	0.000	02:42	
Subject Sample	# 2	02:45	No Sample Provided
Breath Volume	0.000	02:45	
Air Blank	0.000	02:46	
Control Test	0.081	02:46	
Air Blank	0.000	02:46	
Diagnostics Check	OK	02:46	

Data Download Date/Time: 04/27/2016 12:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBINSON,CHRISTOPHER,A
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: ROBINSON,CHRISTOPHER
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 03/01/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 03:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:29

SUBJECT'S LAST NAME: BARTLETT

DATE/TIME OF TEST: 03/13/2016 04:35:36

Diagnostics Check	OK	04:27
Air Blank	0.000	04:27
Control Test	0.081	04:27
Air Blank	0.000	04:28
Subject Sample # 1	0.091	04:31
Breath Volume	2.730	04:31
Air Blank	0.000	04:31
Air Blank	0.000	04:33
Subject Sample # 2	0.086	04:34
Breath Volume	1.738	04:34
Air Blank	0.000	04:34
Control Test	0.080	04:35
Air Blank	0.000	04:35
Diagnostics Check	OK	04:35

Data Download Date/Time: 04/27/2016 12:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARNOLD,AUSTIN,B
BREATH TEST OPERATOR AGENCY: DeFuniak Springs PD
ARRESTING OFFICER: ROCKETT,JONATHAN
ARRESTING OFFICER AGENCY: DeFuniak Springs PD
LAST AGENCY INSPECTION: 03/01/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 13:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:25

SUBJECT'S LAST NAME: HILL

DATE/TIME OF TEST: 03/15/2016 14:53:14

Diagnostics Check	OK	14:46
Air Blank	0.000	14:47
Control Test	0.081	14:47
Air Blank	0.000	14:47
Subject Sample # 1	0.300	14:48
Breath Volume	1.687	14:48
Air Blank	0.000	14:49
Air Blank	0.000	14:51
Subject Sample # 2	0.287	14:51
Breath Volume	1.949	14:51
Air Blank	0.000	14:52
Control Test	0.080	14:52
Air Blank	0.000	14:53
Diagnostics Check	OK	14:53

Data Download Date/Time: 04/27/2016 12:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,CURTIS,C
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: BUSH,NA
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 03/01/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: PROB/PAROL
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:40

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 03/24/2016 14:09:35

Diagnostics Check	OK	14:03
Air Blank	0.000	14:03
Control Test	0.082	14:04
Air Blank	0.000	14:04
Subject Sample # 1	0.000	14:05
Breath Volume	3.757	14:05
Air Blank	0.000	14:05
Air Blank	0.000	14:07
Subject Sample # 2	0.000	14:08
Breath Volume	5.253	14:08
Air Blank	0.000	14:08
Control Test	0.080	14:08
Air Blank	0.000	14:09
Diagnostics Check	OK	14:09

Data Download Date/Time: 04/27/2016 12:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PATE,CHARLES,S
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: NA,NA
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 03/01/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: PROB/PAROL
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:28

SUBJECT'S LAST NAME: ;6360101346010391289

DATE/TIME OF TEST: 03/25/2016 13:02:10

Diagnostics Check	OK	12:56
Air Blank	0.000	12:56
Control Test	0.082	12:57
Air Blank	0.000	12:57
Subject Sample # 1	0.000	12:58
Breath Volume	2.613	12:58
Air Blank	0.000	12:58
Air Blank	0.000	13:00
Subject Sample # 2	0.000	13:00
Breath Volume	2.988	13:00
Air Blank	0.000	13:01
Control Test	0.080	13:01
Air Blank	0.000	13:02
Diagnostics Check	OK	13:02

Data Download Date/Time: 04/27/2016 12:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIERNACKI,CHAD,A
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: BIERNACKI,CHAD
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 03/01/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 02:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 03/06/2016 03:40:34

Diagnostics Check		OK	03:33
Air Blank		0.000	03:34
Control Test		0.080	03:34
Air Blank		0.000	03:35
Subject Sample	# 1	0.237	03:35
Breath Volume		2.503	03:35
Air Blank		0.000	03:36
Air Blank		0.000	03:38
Subject Sample	# 2	0.230	03:39
Breath Volume		2.273	03:39
Air Blank		0.000	03:39
Control Test		0.079	03:39
Air Blank		0.000	03:40
Diagnostics Check		OK	03:40

Data Download Date/Time: 04/27/2016 11:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIERNACKI,CHAD,A
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: BIERNACKI,CHAD
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 03/01/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 20:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:26

SUBJECT'S LAST NAME: RICHARDSON

DATE/TIME OF TEST: 03/06/2016 20:57:40

Diagnostics Check	OK	20:46	
Air Blank	0.000	20:46	
Control Test	0.081	20:47	
Air Blank	0.000	20:47	
Subject Sample	# 1	20:50	No Sample Provided
Breath Volume	0.000	20:50	
Air Blank	0.000	20:51	
Air Blank	0.000	20:53	
Subject Sample	# 2	20:56	No Sample Provided
Breath Volume	0.000	20:56	
Air Blank	0.000	20:56	
Control Test	0.081	20:57	
Air Blank	0.000	20:57	
Diagnostics Check	OK	20:57	

Data Download Date/Time: 04/27/2016 11:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCLEOD,JOEL,A
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: MCCOMICK,CHANTEL
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 03/01/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 22:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:47

SUBJECT'S LAST NAME: SPENCER

DATE/TIME OF TEST: 03/09/2016 23:29:31

Diagnostics Check	OK	23:17	
Air Blank	0.000	23:17	
Control Test	0.079	23:18	
Air Blank	0.000	23:18	
Subject Sample	# 1	23:21	No Sample Provided
Breath Volume	0.000	23:21	
Air Blank	0.000	23:22	
Air Blank	0.000	23:24	
Subject Sample	# 2	23:27	No Sample Provided
Breath Volume	0.000	23:27	
Air Blank	0.000	23:27	
Control Test	0.080	23:28	
Air Blank		23:29	Purge Fail
Air Blank	0.000	23:29	

Data Download Date/Time: 04/27/2016 11:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PULLINGS,ALLEN,C
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: JMCCOY,THOMAS
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 03/01/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 01:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:05

SUBJECT'S LAST NAME: SPEAR

DATE/TIME OF TEST: 03/13/2016 15:34:02

Diagnostics Check	OK	15:25	
Air Blank	0.000	15:25	
Control Test	0.079	15:26	
Air Blank	0.000	15:26	
Subject Sample	# 1	15:29	No Sample Provided
Breath Volume	0.000	15:29	
Air Blank	0.000	15:30	
Air Blank	0.000	15:32	
Subject Sample	# 2	15:32	Subject Test Refused
Breath Volume	0.000	15:32	
Air Blank	0.000	15:33	
Control Test	0.079	15:33	
Air Blank	0.000	15:33	
Diagnostics Check	OK	15:34	

Data Download Date/Time: 04/27/2016 11:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RENNA,DAVID,P
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: N/A,N/A
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 03/01/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: PROB/PAROL
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:04

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 03/13/2016 19:32:45

Diagnostics Check	OK	19:26
Air Blank	0.000	19:27
Control Test	0.080	19:27
Air Blank	0.000	19:28
Subject Sample # 1	0.000	19:28
Breath Volume	3.277	19:28
Air Blank	0.000	19:28
Air Blank	0.000	19:30
Subject Sample # 2	0.000	19:31
Breath Volume	3.781	19:31
Air Blank	0.000	19:31
Control Test	0.078	19:32
Air Blank	0.000	19:32
Diagnostics Check	OK	19:32

Data Download Date/Time: 04/27/2016 11:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RENNA,DAVID,P
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: N/A,N/A
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 03/01/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: PROB/PAROL
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:38

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 03/16/2016 22:17:32

Diagnostics Check	OK	22:11
Air Blank	0.000	22:11
Control Test	0.078	22:12
Air Blank	0.000	22:12
Subject Sample # 1	0.000	22:13
Breath Volume	3.148	22:13
Air Blank	0.000	22:13
Air Blank	0.000	22:15
Subject Sample # 2	0.000	22:16
Breath Volume	2.097	22:16
Air Blank	0.000	22:16
Control Test	0.077	22:16
Air Blank	0.000	22:17
Diagnostics Check	OK	22:17

Data Download Date/Time: 04/27/2016 11:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCLEOD,JOEL,A
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: MCLEOD,JOEL
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 03/01/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 20:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:05

SUBJECT'S LAST NAME: GAULTNEY

DATE/TIME OF TEST: 03/19/2016 22:35:16

Diagnostics Check	OK	22:23	
Air Blank	0.000	22:24	
Control Test	0.078	22:24	
Air Blank	0.000	22:25	
Subject Sample	# 1	22:28	No Sample Provided
Breath Volume	0.000	22:28	
Air Blank	0.000	22:28	
Air Blank	0.000	22:30	
Subject Sample	# 2	22:33	No Sample Provided
Breath Volume	0.000	22:33	
Air Blank	0.000	22:34	
Control Test	0.079	22:34	
Air Blank	0.000	22:35	
Diagnostics Check	OK	22:35	

Data Download Date/Time: 04/27/2016 11:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCLEOD,JOEL,A
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: MCCORMICK,CHANTEL
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 03/01/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 02:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: LORD

DATE/TIME OF TEST: 03/20/2016 03:22:41

Diagnostics Check	OK	03:19	
Air Blank	0.000	03:19	
Control Test	0.079	03:20	
Air Blank	0.000	03:20	
Subject Sample	# 1	03:21	Subject Test Refused
Breath Volume	0.000	03:21	
Air Blank	0.000	03:21	
Control Test	0.080	03:22	
Air Blank	0.000	03:22	
Diagnostics Check	OK	03:22	

Data Download Date/Time: 04/27/2016 11:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAWLES,COREY,M
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: RAWLES,COREY
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 03/01/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 22:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: MCGAHA

DATE/TIME OF TEST: 03/20/2016 23:24:53

Diagnostics Check	OK	23:18
Air Blank	0.000	23:18
Control Test	0.080	23:19
Air Blank	0.000	23:19
Subject Sample # 1	0.135	23:20
Breath Volume	1.757	23:20
Air Blank	0.000	23:21
Air Blank	0.000	23:22
Subject Sample # 2	0.133	23:23
Breath Volume	1.281	23:23
Air Blank	0.000	23:23
Control Test	0.080	23:24
Air Blank	0.000	23:24
Diagnostics Check	OK	23:24

Data Download Date/Time: 04/27/2016 11:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCLEOD,JOEL,A
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: MCLEOD,JOEL
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 03/01/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 04:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:47

SUBJECT'S LAST NAME: RILEY

DATE/TIME OF TEST: 03/25/2016 05:37:18

Diagnostics Check	OK	05:30
Air Blank	0.000	05:31
Control Test	0.079	05:31
Air Blank	0.000	05:32
Subject Sample # 1	0.181	05:32
Breath Volume	2.144	05:32
Air Blank	0.000	05:33
Air Blank	0.000	05:35
Subject Sample # 2	0.175	05:35
Breath Volume	1.753	05:35
Air Blank	0.000	05:36
Control Test	0.079	05:36
Air Blank	0.000	05:37
Diagnostics Check	OK	05:37

Data Download Date/Time: 04/27/2016 11:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PELLITTERI,FELICIA,M
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: PELLITTERI,FELICIA
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 03/01/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: PROB/PAROL
ARREST TIME: 15:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:00

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 03/30/2016 15:39:53

Diagnostics Check	OK	15:33
Air Blank	0.000	15:34
Control Test	0.079	15:34
Air Blank	0.000	15:35
Subject Sample # 1	0.000	15:35
Breath Volume	4.699	15:35
Air Blank	0.000	15:36
Air Blank	0.000	15:38
Subject Sample # 2	0.000	15:38
Breath Volume	2.289	15:38
Air Blank	0.000	15:38
Control Test	0.079	15:39
Air Blank	0.000	15:39
Diagnostics Check	OK	15:39

Data Download Date/Time: 04/27/2016 11:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001320

INSTRUMENT REGISTERED TO: OKEECHOBEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAJERE,ARIC,
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: COLLIER,MARCUS
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 03/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1686121
DRY GAS EXPIRY DATE: 05/21/2017
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:17

SUBJECT'S LAST NAME: LARSON

DATE/TIME OF TEST: 03/19/2016 01:40:33

Diagnostics Check	OK	01:37	
Air Blank	0.000	01:38	
Control Test	0.079	01:38	
Air Blank	0.000	01:39	
Subject Sample	# 1	01:39	Subject Test Refused
Breath Volume	0.000	01:39	
Air Blank	0.000	01:39	
Control Test	0.079	01:39	
Air Blank	0.000	01:40	
Diagnostics Check	OK	01:40	

Data Download Date/Time: 04/12/2016 15:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: IZQUIERDO,ROBERT,W
BREATH TEST OPERATOR AGENCY: FHP Troop K
ARRESTING OFFICER: TILLET,JON
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 02/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1686121
DRY GAS EXPIRY DATE: 05/21/2017
VIOLATION CODE: DUI
ARREST TIME: 05:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:47

SUBJECT'S LAST NAME: SWINNEY

DATE/TIME OF TEST: 03/03/2016 07:36:48

Diagnostics Check	OK	07:29
Air Blank	0.000	07:30
Control Test	0.080	07:30
Air Blank	0.000	07:31
Subject Sample # 1	0.126	07:31
Breath Volume	1.308	07:31
Air Blank	0.000	07:32
Air Blank	0.000	07:34
Subject Sample # 2	0.141	07:35
Breath Volume	2.015	07:35
Air Blank	0.000	07:35
Control Test	0.078	07:36
Air Blank	0.000	07:36
Diagnostics Check	OK	07:36

Data Download Date/Time: 03/14/2016 12:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POLLOCK,STEVEN,D
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: POLLOCK,STEVEN
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 02/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1686121
DRY GAS EXPIRY DATE: 05/21/2017
VIOLATION CODE: DUI
ARREST TIME: 10:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:41

SUBJECT'S LAST NAME: HALL

DATE/TIME OF TEST: 03/04/2016 11:04:00

Diagnostics Check	OK	11:01	
Air Blank	0.000	11:01	
Control Test	0.079	11:01	
Air Blank	0.000	11:02	
Subject Sample	# 1	11:02	Subject Test Refused
Breath Volume	0.000	11:02	
Air Blank	0.000	11:03	
Control Test	0.079	11:03	
Air Blank	0.000	11:03	
Diagnostics Check	OK	11:04	

Data Download Date/Time: 03/14/2016 12:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GITTO,JUSTIN,J
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: GITTO,JUSTIN
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 02/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1686121
DRY GAS EXPIRY DATE: 05/21/2017
VIOLATION CODE: DUI
ARREST TIME: 21:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:35

SUBJECT'S LAST NAME: MALATESTA

DATE/TIME OF TEST: 03/04/2016 23:38:42

Diagnostics Check	OK	23:32
Air Blank	0.000	23:32
Control Test	0.081	23:32
Air Blank	0.000	23:33
Subject Sample # 1	0.049	23:34
Breath Volume	3.117	23:34
Air Blank	0.000	23:34
Air Blank	0.000	23:36
Subject Sample # 2	0.050	23:37
Breath Volume	3.835	23:37
Air Blank	0.000	23:37
Control Test	0.079	23:38
Air Blank	0.000	23:38
Diagnostics Check	OK	23:38

Data Download Date/Time: 03/14/2016 12:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POLLOCK,STEVEN,D
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: COLLIER,MARCUS
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 02/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1686121
DRY GAS EXPIRY DATE: 05/21/2017
VIOLATION CODE: DUI
ARREST TIME: 04:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:13

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 03/06/2016 05:48:10

Diagnostics Check	OK	05:33	
Air Blank	0.000	05:33	
Control Test	0.081	05:34	
Air Blank	0.000	05:34	
Subject Sample	# 1 0.162	05:35	
Breath Volume	1.113	05:35	
Air Blank	0.000	05:36	
Air Blank	0.000	05:38	
Subject Sample	# 2 0.139	05:41	Volume Not Met
Breath Volume	0.101	05:41	
Air Blank	0.000	05:42	
Air Blank	0.000	05:44	
Subject Sample	# 3 0.156	05:46	
Breath Volume	1.964	05:46	
Air Blank	0.000	05:47	
Control Test	0.079	05:47	
Air Blank	0.000	05:48	
Diagnostics Check	OK	05:48	

Data Download Date/Time: 03/14/2016 12:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POLLOCK,STEVEN,D
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: POLLOCK,STEVEN
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 02/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1686121
DRY GAS EXPIRY DATE: 05/21/2017
VIOLATION CODE: DUI
ARREST TIME: 15:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:17

SUBJECT'S LAST NAME: SNOWDEN

DATE/TIME OF TEST: 03/09/2016 15:43:59

Diagnostics Check	OK	15:37
Air Blank	0.000	15:38
Control Test	0.080	15:38
Air Blank	0.000	15:38
Subject Sample # 1	0.318	15:39
Breath Volume	2.382	15:39
Air Blank	0.000	15:40
Air Blank	0.000	15:41
Subject Sample # 2	0.314	15:42
Breath Volume	2.332	15:42
Air Blank	0.000	15:43
Control Test	0.079	15:43
Air Blank	0.000	15:43
Diagnostics Check	OK	15:43

Data Download Date/Time: 03/14/2016 12:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POLLOCK,STEVEN,D
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: JORDAN,MICHAEL
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 02/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1686121
DRY GAS EXPIRY DATE: 05/21/2017
VIOLATION CODE: DUI
ARREST TIME: 14:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:38

SUBJECT'S LAST NAME: BERRY

DATE/TIME OF TEST: 03/10/2016 15:01:10

Diagnostics Check	OK	14:58	
Air Blank	0.000	14:58	
Control Test	0.080	14:59	
Air Blank	0.000	14:59	
Subject Sample	# 1	14:59	Subject Test Refused
Breath Volume	0.000	14:59	
Air Blank	0.000	15:00	
Control Test	0.079	15:00	
Air Blank	0.000	15:01	
Diagnostics Check	OK	15:01	

Data Download Date/Time: 03/14/2016 12:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POLLOCK,STEVEN,D
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: HOLROYD,BRYAN
ARRESTING OFFICER AGENCY: Okeechobee PD
LAST AGENCY INSPECTION: 02/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1686121
DRY GAS EXPIRY DATE: 05/21/2017
VIOLATION CODE: DUI
ARREST TIME: 13:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:11

SUBJECT'S LAST NAME: SAVAGE

DATE/TIME OF TEST: 03/11/2016 14:37:37

Diagnostics Check	OK	14:31
Air Blank	0.000	14:32
Control Test	0.079	14:32
Air Blank	0.000	14:32
Subject Sample # 1	0.227	14:33
Breath Volume	3.558	14:33
Air Blank	0.000	14:33
Air Blank	0.000	14:35
Subject Sample # 2	0.231	14:36
Breath Volume	3.046	14:36
Air Blank	0.000	14:36
Control Test	0.079	14:37
Air Blank	0.000	14:37
Diagnostics Check	OK	14:37

Data Download Date/Time: 03/14/2016 12:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PETO,JOSHUA,W
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: PETO,JOSHUA
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 02/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:56

SUBJECT'S LAST NAME: WHEELER

DATE/TIME OF TEST: 03/08/2016 01:35:54

Diagnostics Check	OK	01:29
Air Blank	0.000	01:29
Control Test	0.079	01:30
Air Blank	0.000	01:30
Subject Sample # 1	0.143	01:31
Breath Volume	4.121	01:31
Air Blank	0.000	01:31
Air Blank	0.000	01:33
Subject Sample # 2	0.145	01:34
Breath Volume	3.847	01:34
Air Blank	0.000	01:35
Control Test	0.079	01:35
Air Blank	0.000	01:35
Diagnostics Check	OK	01:35

Data Download Date/Time: 03/26/2016 10:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OJEDA,NELSON,J
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: OJEDA,NELSON
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 02/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: BLAISE

DATE/TIME OF TEST: 03/09/2016 01:09:48

Diagnostics Check		OK	01:02
Air Blank		0.000	01:03
Control Test		0.079	01:03
Air Blank		0.000	01:04
Subject Sample	# 1	0.204	01:05
Breath Volume		1.273	01:05
Air Blank		0.000	01:05
Air Blank		0.000	01:07
Subject Sample	# 2	0.201	01:08
Breath Volume		1.117	01:08
Air Blank		0.000	01:08
Control Test		0.079	01:09
Air Blank		0.000	01:09
Diagnostics Check		OK	01:09

Data Download Date/Time: 03/26/2016 10:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEWIS,JEREMY,A
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: HARDISON,AISHA
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 02/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 12:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:08

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 03/17/2016 01:36:31

Diagnostics Check	OK	01:28
Air Blank	0.000	01:29
Control Test	0.078	01:29
Air Blank	0.000	01:29
Subject Sample # 1	0.148	01:31
Breath Volume	1.183	01:31
Air Blank	0.000	01:32
Air Blank	0.000	01:34
Subject Sample # 2	0.162	01:34
Breath Volume	1.417	01:34
Air Blank	0.000	01:35
Control Test	0.078	01:35
Air Blank	0.000	01:36
Diagnostics Check	OK	01:36

Data Download Date/Time: 03/26/2016 10:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SISTARELLI,NICHOLAS,J
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: MCINEREY,JESSE
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 02/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: FILEY

DATE/TIME OF TEST: 03/22/2016 02:00:45

Diagnostics Check	OK	01:54
Air Blank	0.000	01:55
Control Test	0.080	01:55
Air Blank	0.000	01:55
Subject Sample # 1	0.106	01:56
Breath Volume	2.386	01:56
Air Blank	0.000	01:57
Air Blank	0.000	01:58
Subject Sample # 2	0.105	01:59
Breath Volume	2.187	01:59
Air Blank	0.000	01:59
Control Test	0.080	02:00
Air Blank	0.000	02:00
Diagnostics Check	OK	02:00

Data Download Date/Time: 03/26/2016 10:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YORK, JIMMY, A
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: LANHAM, MICHAEL
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 02/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 237365
DRY GAS EXPIRY DATE: 01/04/2019
VIOLATION CODE: DUI
ARREST TIME: 03:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:00

SUBJECT'S LAST NAME: HUBBARD

DATE/TIME OF TEST: 03/01/2016 06:27:26

Diagnostics Check	OK	06:20
Air Blank	0.000	06:21
Control Test	0.081	06:21
Air Blank	0.000	06:22
Subject Sample # 1	0.131	06:22
Breath Volume	2.945	06:22
Air Blank	0.000	06:23
Air Blank	0.000	06:25
Subject Sample # 2	0.132	06:25
Breath Volume	3.246	06:25
Air Blank	0.000	06:26
Control Test	0.082	06:26
Air Blank	0.000	06:27
Diagnostics Check	OK	06:27

Data Download Date/Time: 03/25/2016 05:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STOVALL,C,J
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: RUIZ,M
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 02/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 237365
DRY GAS EXPIRY DATE: 01/04/2019
VIOLATION CODE: DUI
ARREST TIME: 20:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: WATERSTRIPE

DATE/TIME OF TEST: 03/01/2016 21:53:53

Diagnostics Check		OK	21:42	
Air Blank		0.000	21:42	
Control Test		0.081	21:43	
Air Blank		0.000	21:43	
Subject Sample	# 1	0.200	21:46	Volume Not Met
Breath Volume		0.242	21:46	
Air Blank		0.000	21:47	
Air Blank		0.000	21:49	
Subject Sample	# 2	0.203	21:52	Volume Not Met
Breath Volume		0.128	21:52	
Air Blank		0.000	21:52	
Control Test		0.079	21:53	
Air Blank		0.000	21:53	
Diagnostics Check		OK	21:53	

Data Download Date/Time: 03/25/2016 05:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STOVALL,C,J
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: RUIZ,M
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 02/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 237365
DRY GAS EXPIRY DATE: 01/04/2019
VIOLATION CODE: DUI
ARREST TIME: 20:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: WATERSTRIPE

DATE/TIME OF TEST: 03/01/2016 22:07:30

Diagnostics Check		OK	21:55	
Air Blank		0.000	21:56	
Control Test		0.079	21:56	
Air Blank		0.000	21:57	
Subject Sample	# 1	0.183	22:00	Volume Not Met
Breath Volume		0.789	22:00	
Air Blank		0.000	22:00	
Air Blank		0.000	22:02	
Subject Sample	# 2	0.201	22:05	Volume Not Met
Breath Volume		0.449	22:05	
Air Blank		0.000	22:06	
Control Test		0.079	22:06	
Air Blank		0.000	22:07	
Diagnostics Check		OK	22:07	

Data Download Date/Time: 03/25/2016 05:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EGGERS,RYAN,K
BREATH TEST OPERATOR AGENCY: Vero Beach PD
ARRESTING OFFICER: EGGERS,RYAN
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 02/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 237365
DRY GAS EXPIRY DATE: 01/04/2019
VIOLATION CODE: DUI
ARREST TIME: 01:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:39

SUBJECT'S LAST NAME: GRIFFIN

DATE/TIME OF TEST: 03/02/2016 02:24:10

Diagnostics Check	OK	02:23	
Air Blank	0.000	02:23	
		02:24	RFI Detect
Air Blank		02:24	RFI Detect

Data Download Date/Time: 03/25/2016 05:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EGGERS,RYAN,K
BREATH TEST OPERATOR AGENCY: Vero Beach PD
ARRESTING OFFICER: EGGERS,RYAN
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 02/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 237365
DRY GAS EXPIRY DATE: 01/04/2019
VIOLATION CODE: DUI
ARREST TIME: 01:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:39

SUBJECT'S LAST NAME: GRIFFIN

DATE/TIME OF TEST: 03/02/2016 02:31:44

Diagnostics Check	OK	02:25
Air Blank	0.000	02:26
Control Test	0.080	02:26
Air Blank	0.000	02:26
Subject Sample # 1	0.222	02:27
Breath Volume	3.238	02:27
Air Blank	0.000	02:27
Air Blank	0.000	02:29
Subject Sample # 2	0.214	02:30
Breath Volume	2.562	02:30
Air Blank	0.000	02:30
Control Test	0.079	02:31
Air Blank	0.000	02:31
Diagnostics Check	OK	02:31

Data Download Date/Time: 03/25/2016 05:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSADO,L,M
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: DEAN,A
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 02/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 237365
DRY GAS EXPIRY DATE: 01/04/2019
VIOLATION CODE: DUI
ARREST TIME: 01:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: LEMAY

DATE/TIME OF TEST: 03/05/2016 02:59:40

Diagnostics Check	OK	02:56	
Air Blank	0.000	02:57	
Control Test	0.081	02:57	
Air Blank	0.000	02:58	
Subject Sample	# 1	02:58	Subject Test Refused
Breath Volume	0.000	02:58	
Air Blank	0.000	02:58	
Control Test	0.081	02:59	
Air Blank	0.000	02:59	
Diagnostics Check	OK	02:59	

Data Download Date/Time: 03/25/2016 05:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROLLE,MELINDA,M
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: JAWORSKI,KEVIN
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 02/13/2016

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 237365
DRY GAS EXPIRY DATE: 01/04/2019
VIOLATION CODE: DUI
ARREST TIME: 21:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: DICK

DATE/TIME OF TEST: 03/05/2016 22:08:28

Diagnostics Check	OK	22:00
Air Blank	0.000	22:01
Control Test	0.082	22:01
Air Blank	0.000	22:02
Subject Sample # 1	0.173	22:04
Breath Volume	1.953	22:04
Air Blank	0.000	22:04
Air Blank	0.000	22:06
Subject Sample # 2	0.168	22:06
Breath Volume	1.871	22:06
Air Blank	0.000	22:07
Control Test	0.080	22:07
Air Blank	0.000	22:08
Diagnostics Check	OK	22:08

Data Download Date/Time: 03/25/2016 05:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRIS,JOSHUA,D
BREATH TEST OPERATOR AGENCY: Vero Beach PD
ARRESTING OFFICER: EDER,KYLE
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 02/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 237365
DRY GAS EXPIRY DATE: 01/04/2019
VIOLATION CODE: DUI
ARREST TIME: 04:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:45

SUBJECT'S LAST NAME: VELAZQUEZ

DATE/TIME OF TEST: 03/06/2016 05:13:48

Diagnostics Check	OK	05:10	
Air Blank	0.000	05:11	
Control Test	0.081	05:11	
Air Blank	0.000	05:12	
Subject Sample	# 1	05:12	Subject Test Refused
Breath Volume	0.000	05:12	
Air Blank	0.000	05:12	
Control Test	0.081	05:13	
Air Blank	0.000	05:13	
Diagnostics Check	OK	05:13	

Data Download Date/Time: 03/25/2016 05:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VECCHIONE, ADAM, J
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: DEAN, ADAM
ARRESTING OFFICER AGENCY: Fellsmere PD
LAST AGENCY INSPECTION: 02/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 237365
DRY GAS EXPIRY DATE: 01/04/2019
VIOLATION CODE: DUI
ARREST TIME: 20:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: DECKER

DATE/TIME OF TEST: 03/06/2016 22:24:51

Diagnostics Check	OK	22:21	
Air Blank	0.000	22:22	
Control Test	0.082	22:22	
Air Blank	0.000	22:23	
Subject Sample	# 1	22:23	Subject Test Refused
Breath Volume	0.000	22:23	
Air Blank	0.000	22:23	
Control Test	0.081	22:24	
Air Blank	0.000	22:24	
Diagnostics Check	OK	22:24	

Data Download Date/Time: 03/25/2016 05:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROLLE,MELINDA,M
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: IOVINO,ALBERT
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 02/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 237365
DRY GAS EXPIRY DATE: 01/04/2019
VIOLATION CODE: DUI
ARREST TIME: 18:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:05

SUBJECT'S LAST NAME: OSTERHOUT

DATE/TIME OF TEST: 03/07/2016 19:32:01

Diagnostics Check	OK	19:25
Air Blank	0.000	19:26
Control Test	0.082	19:26
Air Blank	0.000	19:27
Subject Sample # 1	0.170	19:27
Breath Volume	1.437	19:27
Air Blank	0.000	19:28
Air Blank	0.000	19:30
Subject Sample # 2	0.172	19:30
Breath Volume	1.281	19:30
Air Blank	0.000	19:31
Control Test	0.082	19:31
Air Blank	0.000	19:31
Diagnostics Check	OK	19:32

Data Download Date/Time: 03/25/2016 05:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GINSBURG, DOUGLAS, J
BREATH TEST OPERATOR AGENCY: Vero Beach PD
ARRESTING OFFICER: GINSBURG, DOUGLAS
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 02/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 237365
DRY GAS EXPIRY DATE: 01/04/2019
VIOLATION CODE: DUI
ARREST TIME: 03:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:22

SUBJECT'S LAST NAME: NADOLNY

DATE/TIME OF TEST: 03/08/2016 03:49:30

Diagnostics Check		OK	03:43
Air Blank		0.000	03:43
Control Test		0.082	03:43
Air Blank		0.000	03:44
Subject Sample	# 1	0.214	03:44
Breath Volume		4.316	03:44
Air Blank		0.000	03:45
Air Blank		0.000	03:47
Subject Sample	# 2	0.202	03:47
Breath Volume		3.878	03:47
Air Blank		0.000	03:48
Control Test		0.081	03:48
Air Blank		0.000	03:49
Diagnostics Check		OK	03:49

Data Download Date/Time: 03/25/2016 05:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GINSBURG, DOUGLAS, J
BREATH TEST OPERATOR AGENCY: Vero Beach PD
ARRESTING OFFICER: GINSBURG, DOUGLAS
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 02/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 237365
DRY GAS EXPIRY DATE: 01/04/2019
VIOLATION CODE: DUI
ARREST TIME: 02:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:03

SUBJECT'S LAST NAME: CROSBY

DATE/TIME OF TEST: 03/09/2016 02:33:39

Diagnostics Check	OK	02:27
Air Blank	0.000	02:28
Control Test	0.080	02:28
Air Blank	0.000	02:28
Subject Sample # 1	0.087	02:29
Breath Volume	3.820	02:29
Air Blank	0.000	02:29
Air Blank	0.000	02:31
Subject Sample # 2	0.087	02:32
Breath Volume	3.609	02:32
Air Blank	0.000	02:32
Control Test	0.081	02:33
Air Blank	0.000	02:33
Diagnostics Check	OK	02:33

Data Download Date/Time: 03/25/2016 05:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOLSKI,JEFFREY,R
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: JAWORSKI,KEVIN
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 02/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 237365
DRY GAS EXPIRY DATE: 01/04/2019
VIOLATION CODE: DUI
ARREST TIME: 21:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: MITCHELL

DATE/TIME OF TEST: 03/11/2016 22:06:21

Diagnostics Check	OK	21:59
Air Blank	0.000	22:00
Control Test	0.081	22:00
Air Blank	0.000	22:01
Subject Sample # 1	0.285	22:01
Breath Volume	3.093	22:01
Air Blank	0.000	22:02
Air Blank	0.000	22:04
Subject Sample # 2	0.265	22:04
Breath Volume	2.300	22:04
Air Blank	0.000	22:05
Control Test	0.081	22:05
Air Blank	0.000	22:06
Diagnostics Check	OK	22:06

Data Download Date/Time: 03/25/2016 05:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROLLE,MELINDA,M
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: JAWORSKI,KEVIN
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 02/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 237365
DRY GAS EXPIRY DATE: 01/04/2019
VIOLATION CODE: DUI
ARREST TIME: 20:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:15

SUBJECT'S LAST NAME: SIMILIEN

DATE/TIME OF TEST: 03/12/2016 21:57:45

Diagnostics Check	OK	21:56	
Air Blank	0.000	21:56	
Control Test	0.081	21:57	
Air Blank		21:57	RFI Detect
Air Blank	0.000	21:57	

Data Download Date/Time: 03/25/2016 05:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROLLE,MELINDA,M
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: JAWORSKI,KEVIN
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 02/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 237365
DRY GAS EXPIRY DATE: 01/04/2019
VIOLATION CODE: DUI
ARREST TIME: 20:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:15

SUBJECT'S LAST NAME: SIMILIEN

DATE/TIME OF TEST: 03/12/2016 22:04:15

Diagnostics Check	OK	21:59	
Air Blank	0.000	22:00	
Control Test	0.081	22:00	
Air Blank	0.000	22:01	
Subject Sample	# 1	22:02	Subject Test Refused
Breath Volume	0.199	22:02	
Air Blank	0.000	22:03	
Control Test	0.081	22:03	
Air Blank	0.000	22:04	
Diagnostics Check	OK	22:04	

Data Download Date/Time: 03/25/2016 05:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GINSBURG, DOUGLAS, J
BREATH TEST OPERATOR AGENCY: Vero Beach PD
ARRESTING OFFICER: JOERGER, SHANE
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 02/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 237365
DRY GAS EXPIRY DATE: 01/04/2019
VIOLATION CODE: DUI
ARREST TIME: 00:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:41

SUBJECT'S LAST NAME: MCKINNON

DATE/TIME OF TEST: 03/13/2016 01:15:39

Diagnostics Check	OK	01:12	
Air Blank	0.000	01:13	
Control Test	0.081	01:13	
Air Blank	0.000	01:14	
Subject Sample	# 1	01:14	Subject Test Refused
Breath Volume	0.000	01:14	
Air Blank	0.000	01:14	
Control Test	0.081	01:15	
Air Blank	0.000	01:15	
Diagnostics Check	OK	01:15	

Data Download Date/Time: 03/25/2016 05:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TANGEL, MATTHEW, A
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: SMITH, COLBY
ARRESTING OFFICER AGENCY: Sebastian PD
LAST AGENCY INSPECTION: 02/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 237365
DRY GAS EXPIRY DATE: 01/04/2019
VIOLATION CODE: DUI
ARREST TIME: 00:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:54

SUBJECT'S LAST NAME: CORNWELL

DATE/TIME OF TEST: 03/14/2016 01:27:33

Diagnostics Check		OK	01:15	
Air Blank		0.000	01:15	
Control Test		0.081	01:15	
Air Blank		0.000	01:16	
Subject Sample	# 1	0.282	01:19	Volume Not Met
Breath Volume		0.980	01:19	
Air Blank		0.000	01:20	
Air Blank		0.000	01:22	
Subject Sample	# 2	0.283	01:23	
Breath Volume		1.449	01:23	
Air Blank		0.000	01:23	
Air Blank		0.000	01:25	
Subject Sample	# 3	0.279	01:25	
Breath Volume		1.398	01:25	
Air Blank		0.000	01:26	
Control Test		0.079	01:26	
Air Blank		0.000	01:27	
Diagnostics Check		OK	01:27	

Data Download Date/Time: 03/25/2016 05:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOLSKI,JEFFREY,R
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: FLORES,DAVID
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 02/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 237365
DRY GAS EXPIRY DATE: 01/04/2019
VIOLATION CODE: DUI
ARREST TIME: 21:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: HANNA

DATE/TIME OF TEST: 03/14/2016 22:37:32

Diagnostics Check	OK	22:31
Air Blank	0.000	22:31
Control Test	0.081	22:32
Air Blank	0.000	22:32
Subject Sample # 1	0.201	22:33
Breath Volume	1.593	22:33
Air Blank	0.000	22:33
Air Blank	0.000	22:35
Subject Sample # 2	0.218	22:36
Breath Volume	1.960	22:36
Air Blank	0.000	22:36
Control Test	0.080	22:36
Air Blank	0.000	22:37
Diagnostics Check	OK	22:37

Data Download Date/Time: 03/25/2016 05:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GILBERT,TRESA,M
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: SHARPETA,CHAD
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 02/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 237365
DRY GAS EXPIRY DATE: 01/04/2019
VIOLATION CODE: DUI
ARREST TIME: 23:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: VANSUSTEREN

DATE/TIME OF TEST: 03/16/2016 00:49:30

Diagnostics Check	OK	00:46	
Air Blank	0.000	00:47	
Control Test	0.081	00:47	
Air Blank	0.000	00:47	
Subject Sample	# 1	00:48	Subject Test Refused
Breath Volume	0.000	00:48	
Air Blank	0.000	00:48	
Control Test	0.081	00:48	
Air Blank	0.000	00:49	
Diagnostics Check	OK	00:49	

Data Download Date/Time: 03/25/2016 05:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EGGERS,RYAN,K
BREATH TEST OPERATOR AGENCY: Vero Beach PD
ARRESTING OFFICER: EGGERS,RYAN
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 02/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 237365
DRY GAS EXPIRY DATE: 01/04/2019
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 03/18/2016 01:54:51

Diagnostics Check	OK	01:52	
Air Blank	0.000	01:52	
Control Test	0.081	01:52	
Air Blank	0.000	01:53	
Subject Sample	# 1	01:53	Subject Test Refused
Breath Volume	0.000	01:53	
Air Blank	0.000	01:53	
Control Test	0.081	01:54	
Air Blank	0.000	01:54	
Diagnostics Check	OK	01:54	

Data Download Date/Time: 03/25/2016 05:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STOVALL,C,J
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: JAWORSKI,K
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 02/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 237365
DRY GAS EXPIRY DATE: 01/04/2019
VIOLATION CODE: DUI
ARREST TIME: 21:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 03/19/2016 22:37:26

Diagnostics Check	OK	22:31
Air Blank	0.000	22:31
Control Test	0.082	22:31
Air Blank	0.000	22:32
Subject Sample # 1	0.286	22:32
Breath Volume	2.148	22:32
Air Blank	0.000	22:33
Air Blank	0.000	22:35
Subject Sample # 2	0.293	22:35
Breath Volume	2.406	22:35
Air Blank	0.000	22:36
Control Test	0.081	22:36
Air Blank	0.000	22:37
Diagnostics Check	OK	22:37

Data Download Date/Time: 03/25/2016 05:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STOVALL,C,J
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: DEAN,A
ARRESTING OFFICER AGENCY: Fellsmere PD
LAST AGENCY INSPECTION: 02/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 237365
DRY GAS EXPIRY DATE: 01/04/2019
VIOLATION CODE: DUI
ARREST TIME: 21:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: STINSON

DATE/TIME OF TEST: 03/19/2016 23:06:05

Diagnostics Check	OK	23:03	
Air Blank	0.000	23:03	
Control Test	0.081	23:04	
Air Blank	0.000	23:04	
Subject Sample	# 1	23:04	Subject Test Refused
Breath Volume	0.000	23:04	
Air Blank	0.000	23:05	
Control Test	0.082	23:05	
Air Blank	0.000	23:06	
Diagnostics Check	OK	23:06	

Data Download Date/Time: 03/25/2016 05:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSADO,L,M
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: EARMAN,S
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 02/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 237365
DRY GAS EXPIRY DATE: 01/04/2019
VIOLATION CODE: DUI
ARREST TIME: 01:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:23

SUBJECT'S LAST NAME: FESTIAN

DATE/TIME OF TEST: 03/20/2016 02:55:43

Diagnostics Check	OK	02:52	
Air Blank	0.000	02:53	
Control Test	0.081	02:53	
Air Blank	0.000	02:54	
Subject Sample	# 1	02:54	Subject Test Refused
Breath Volume	0.000	02:54	
Air Blank	0.000	02:54	
Control Test	0.081	02:55	
Air Blank	0.000	02:55	
Diagnostics Check	OK	02:55	

Data Download Date/Time: 03/25/2016 05:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STOVALL,C,J
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: BUFFINGTON,B
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 02/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 237365
DRY GAS EXPIRY DATE: 01/04/2019
VIOLATION CODE: DUI
ARREST TIME: 17:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:00

SUBJECT'S LAST NAME: BLACKWILL

DATE/TIME OF TEST: 03/21/2016 18:38:20

Diagnostics Check	OK	18:35	
Air Blank	0.000	18:36	
Control Test	0.082	18:36	
Air Blank	0.000	18:36	
Subject Sample	# 1	18:36	Subject Test Refused
Breath Volume	0.000	18:36	
Air Blank	0.000	18:37	
Control Test	0.082	18:37	
Air Blank	0.000	18:38	
Diagnostics Check	OK	18:38	

Data Download Date/Time: 03/25/2016 05:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EGGERS,RYAN,K
BREATH TEST OPERATOR AGENCY: Vero Beach PD
ARRESTING OFFICER: EGGERS,RYAN
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 02/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 237365
DRY GAS EXPIRY DATE: 01/04/2019
VIOLATION CODE: DUI
ARREST TIME: 21:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: DIOS

DATE/TIME OF TEST: 03/21/2016 21:49:50

Diagnostics Check	OK	21:42
Air Blank	0.000	21:42
Control Test	0.082	21:43
Air Blank	0.000	21:43
Subject Sample # 1	0.090	21:44
Breath Volume	1.480	21:44
Air Blank	0.000	21:45
Air Blank	0.000	21:46
Subject Sample # 2	0.081	21:48
Breath Volume	1.476	21:48
Air Blank	0.000	21:48
Control Test	0.082	21:49
Air Blank	0.000	21:49
Diagnostics Check	OK	21:49

Data Download Date/Time: 03/25/2016 05:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EGGERS,RYAN,K
BREATH TEST OPERATOR AGENCY: Vero Beach PD
ARRESTING OFFICER: EGGERS,RYAN
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 03/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 237365
DRY GAS EXPIRY DATE: 01/04/2019
VIOLATION CODE: DUI
ARREST TIME: 23:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: OPGAARD

DATE/TIME OF TEST: 03/25/2016 23:47:05

Diagnostics Check	OK	23:40
Air Blank	0.000	23:41
Control Test	0.081	23:41
Air Blank	0.000	23:42
Subject Sample # 1	0.128	23:42
Breath Volume	1.492	23:42
Air Blank	0.000	23:43
Air Blank	0.000	23:44
Subject Sample # 2	0.145	23:45
Breath Volume	2.105	23:45
Air Blank	0.000	23:46
Control Test	0.081	23:46
Air Blank	0.000	23:47
Diagnostics Check	OK	23:47

Data Download Date/Time: 04/27/2016 04:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TANGEL, MATTHEW, A
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: JAWORSKI, KEVIN
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 03/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 237365
DRY GAS EXPIRY DATE: 01/04/2019
VIOLATION CODE: DUI
ARREST TIME: 00:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:42

SUBJECT'S LAST NAME: FOX

DATE/TIME OF TEST: 03/26/2016 01:13:44

Diagnostics Check	OK	01:10	
Air Blank	0.000	01:11	
Control Test	0.081	01:11	
Air Blank	0.000	01:12	
Subject Sample	# 1	01:12	Subject Test Refused
Breath Volume	0.000	01:12	
Air Blank	0.000	01:12	
Control Test	0.080	01:13	
Air Blank	0.000	01:13	
Diagnostics Check	OK	01:13	

Data Download Date/Time: 04/27/2016 04:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GINSBURG, DOUGLAS, J
BREATH TEST OPERATOR AGENCY: Vero Beach PD
ARRESTING OFFICER: CROWLEY, SEAN
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 03/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 237365
DRY GAS EXPIRY DATE: 01/04/2019
VIOLATION CODE: DUI
ARREST TIME: 04:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:38

SUBJECT'S LAST NAME: FISH JR

DATE/TIME OF TEST: 03/26/2016 05:09:57

Diagnostics Check	OK	05:00
Air Blank	0.000	05:01
Control Test	0.081	05:01
Air Blank	0.000	05:01
Subject Sample # 1	0.189	05:02
Breath Volume	3.605	05:02
Air Blank	0.000	05:03
Air Blank	0.000	05:04
Subject Sample # 2	0.156	05:05
Breath Volume	1.910	05:05
Air Blank	0.000	05:05
Air Blank	0.000	05:07
Subject Sample # 3	0.194	05:08
Breath Volume	4.621	05:08
Air Blank	0.000	05:09
Control Test	0.079	05:09
Air Blank	0.000	05:09
Diagnostics Check	OK	05:09

Data Download Date/Time: 04/27/2016 04:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STOVALL,C,J
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: CARROLL,E
ARRESTING OFFICER AGENCY: Sebastian PD
LAST AGENCY INSPECTION: 03/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 237365
DRY GAS EXPIRY DATE: 01/04/2019
VIOLATION CODE: DUI
ARREST TIME: 20:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: SPANN

DATE/TIME OF TEST: 03/26/2016 21:59:08

Diagnostics Check	OK	21:56	
Air Blank	0.000	21:56	
Control Test	0.081	21:57	
Air Blank	0.000	21:57	
Subject Sample	# 1	21:57	Subject Test Refused
Breath Volume	0.000	21:57	
Air Blank	0.000	21:58	
Control Test	0.080	21:58	
Air Blank	0.000	21:59	
Diagnostics Check	OK	21:59	

Data Download Date/Time: 04/27/2016 04:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TANGEL, MATTHEW, A
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: WILLIS, JEREMIAH
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 03/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 237365
DRY GAS EXPIRY DATE: 01/04/2019
VIOLATION CODE: DUI
ARREST TIME: 01:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: MELNICK

DATE/TIME OF TEST: 03/27/2016 02:27:51

Diagnostics Check	OK	02:25	
Air Blank	0.000	02:25	
Control Test	0.082	02:25	
Air Blank	0.000	02:26	
Subject Sample	# 1	02:26	Subject Test Refused
Breath Volume	0.000	02:26	
Air Blank	0.000	02:26	
Control Test	0.081	02:27	
Air Blank	0.000	02:27	
Diagnostics Check	OK	02:27	

Data Download Date/Time: 04/27/2016 04:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOLSKI,JEFFREY,R
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: JAWORSKI,KEVIN
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 03/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 237365
DRY GAS EXPIRY DATE: 01/04/2019
VIOLATION CODE: DUI
ARREST TIME: 22:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: ARCHER

DATE/TIME OF TEST: 03/30/2016 23:49:29

Diagnostics Check	OK	23:46	
Air Blank	0.000	23:47	
Control Test	0.081	23:47	
Air Blank	0.000	23:47	
Subject Sample	# 1	23:48	Subject Test Refused
Breath Volume	0.000	23:48	
Air Blank	0.000	23:48	
Control Test	0.082	23:48	
Air Blank	0.000	23:49	
Diagnostics Check	OK	23:49	

Data Download Date/Time: 04/27/2016 04:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOLSKI,JEFFREY,R
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: LAWHORN,J
ARRESTING OFFICER AGENCY: Sebastian PD
LAST AGENCY INSPECTION: 03/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 237365
DRY GAS EXPIRY DATE: 01/04/2019
VIOLATION CODE: DUI
ARREST TIME: 18:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:55

SUBJECT'S LAST NAME: CIBOROWSKI

DATE/TIME OF TEST: 03/31/2016 19:22:00

Diagnostics Check	OK	19:16
Air Blank	0.000	19:16
Control Test	0.082	19:16
Air Blank	0.000	19:17
Subject Sample # 1	0.210	19:17
Breath Volume	1.597	19:17
Air Blank	0.000	19:18
Air Blank	0.000	19:20
Subject Sample # 2	0.198	19:20
Breath Volume	1.722	19:20
Air Blank	0.000	19:21
Control Test	0.081	19:21
Air Blank	0.000	19:21
Diagnostics Check	OK	19:22

Data Download Date/Time: 04/27/2016 04:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLINT,SERGEANT,C
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: FLINT,SERGEANT
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 03/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: COLLUM

DATE/TIME OF TEST: 03/26/2016 00:20:32

Diagnostics Check	OK	00:10
Air Blank	0.000	00:11
Control Test	0.077	00:11
Air Blank	0.000	00:11
Subject Sample # 1	0.100	00:12
Breath Volume	1.496	00:12
Air Blank	0.000	00:13
Air Blank	0.000	00:15
Subject Sample # 2	0.124	00:15
Breath Volume	3.882	00:15
Air Blank	0.000	00:16
Air Blank	0.000	00:18
Subject Sample # 3	0.114	00:19
Breath Volume	3.394	00:19
Air Blank	0.000	00:19
Control Test	0.077	00:19
Air Blank	0.000	00:20
Diagnostics Check	OK	00:20

Data Download Date/Time: 04/28/2016 09:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NAMORS,JASON,R
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: ANDERSON,K
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 03/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: CARROLL

DATE/TIME OF TEST: 03/26/2016 23:55:20

Diagnostics Check	OK	23:48
Air Blank	0.000	23:49
Control Test	0.078	23:49
Air Blank	0.000	23:50
Subject Sample # 1	0.096	23:50
Breath Volume	1.386	23:50
Air Blank	0.000	23:51
Air Blank	0.000	23:53
Subject Sample # 2	0.094	23:53
Breath Volume	1.535	23:53
Air Blank	0.000	23:54
Control Test	0.078	23:54
Air Blank	0.000	23:55
Diagnostics Check	OK	23:55

Data Download Date/Time: 04/28/2016 09:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSS, MICHAEL, K
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: KOSTYAL, DEPUTY
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 03/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:49

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 03/27/2016 06:25:08

Diagnostics Check		OK	06:18
Air Blank		0.000	06:18
Control Test		0.077	06:18
Air Blank		0.000	06:19
Subject Sample	# 1	0.128	06:20
Breath Volume		2.265	06:20
Air Blank		0.000	06:21
Air Blank		0.000	06:23
Subject Sample	# 2	0.130	06:23
Breath Volume		2.687	06:23
Air Blank		0.000	06:24
Control Test		0.077	06:24
Air Blank		0.000	06:25
Diagnostics Check		OK	06:25

Data Download Date/Time: 04/28/2016 09:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSS, MICHAEL, K
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: HEYMAN, DEPUTY
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 03/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 06:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:43

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 03/27/2016 07:28:09

Diagnostics Check	OK	07:21
Air Blank	0.000	07:22
Control Test	0.078	07:22
Air Blank	0.000	07:22
Subject Sample # 1	0.126	07:23
Breath Volume	1.316	07:23
Air Blank	0.000	07:24
Air Blank	0.000	07:26
Subject Sample # 2	0.129	07:26
Breath Volume	1.156	07:26
Air Blank	0.000	07:27
Control Test	0.077	07:27
Air Blank	0.000	07:28
Diagnostics Check	OK	07:28

Data Download Date/Time: 04/28/2016 09:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,DANIEL,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: SEATON,DEPUTY
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 03/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:27

SUBJECT'S LAST NAME: PENTZ

DATE/TIME OF TEST: 03/29/2016 03:04:41

Diagnostics Check	OK	02:56
Air Blank	0.000	02:57
Control Test	0.078	02:57
Air Blank	0.000	02:58
Subject Sample # 1	0.219	02:59
Breath Volume	1.332	02:59
Air Blank	0.000	03:00
Air Blank	0.000	03:02
Subject Sample # 2	0.222	03:03
Breath Volume	1.468	03:03
Air Blank	0.000	03:03
Control Test	0.077	03:04
Air Blank	0.000	03:04
Diagnostics Check	OK	03:04

Data Download Date/Time: 04/28/2016 09:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAN ARSDALE,JEFFREY,A
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: VAN ARSDALE,JEFREY
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 03/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:06

SUBJECT'S LAST NAME: CARNEY

DATE/TIME OF TEST: 03/29/2016 22:52:33

Diagnostics Check	OK	22:46
Air Blank	0.000	22:46
Control Test	0.077	22:47
Air Blank	0.000	22:47
Subject Sample # 1	0.064	22:48
Breath Volume	2.320	22:48
Air Blank	0.000	22:48
Air Blank	0.000	22:50
Subject Sample # 2	0.063	22:51
Breath Volume	2.042	22:51
Air Blank	0.000	22:51
Control Test	0.078	22:51
Air Blank	0.000	22:52
Diagnostics Check	OK	22:52

Data Download Date/Time: 04/28/2016 09:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NICK,A.,
BREATH TEST OPERATOR AGENCY: North Port PD
ARRESTING OFFICER: NICK,A
ARRESTING OFFICER AGENCY: North Port PD
LAST AGENCY INSPECTION: 03/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: JANSKI

DATE/TIME OF TEST: 03/30/2016 03:57:52

Diagnostics Check	OK	03:51
Air Blank	0.000	03:52
Control Test	0.077	03:52
Air Blank	0.000	03:52
Subject Sample # 1	0.000	03:53
Breath Volume	1.695	03:53
Air Blank	0.000	03:53
Air Blank	0.000	03:55
Subject Sample # 2	0.000	03:56
Breath Volume	2.414	03:56
Air Blank	0.000	03:56
Control Test	0.076	03:57
Air Blank	0.000	03:57
Diagnostics Check	OK	03:57

Data Download Date/Time: 04/28/2016 09:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FOWLER, TYLER, J
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: FOWLER, TYLER
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 03/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 17:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:14

SUBJECT'S LAST NAME: CLEMENTI

DATE/TIME OF TEST: 03/31/2016 18:49:06

Diagnostics Check	OK	18:41
Air Blank	0.000	18:42
Control Test	0.079	18:42
Air Blank	0.000	18:43
Subject Sample # 1	0.172	18:43
Breath Volume	1.152	18:43
Air Blank	0.000	18:44
Air Blank	0.000	18:46
Subject Sample # 2	0.171	18:47
Breath Volume	1.335	18:47
Air Blank	0.000	18:48
Control Test	0.076	18:48
Air Blank	0.000	18:49
Diagnostics Check	OK	18:49

Data Download Date/Time: 04/28/2016 09:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

INSTRUMENT REGISTERED TO: EUSTIS P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWELL,SHAWNA,M
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: HOWELL,SHAWNA
ARRESTING OFFICER AGENCY: Eustis PD
LAST AGENCY INSPECTION: 03/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1787697
DRY GAS EXPIRY DATE: 01/08/2018
VIOLATION CODE: DUI
ARREST TIME: 02:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:13

SUBJECT'S LAST NAME: BAEZ

DATE/TIME OF TEST: 03/18/2016 03:51:01

Diagnostics Check	OK	03:42	
Air Blank	0.000	03:42	
Control Test	0.079	03:43	
Air Blank	0.000	03:43	
Subject Sample	# 1	03:46	No Sample Provided
Breath Volume	0.000	03:46	
Air Blank	0.000	03:47	
Air Blank	0.000	03:49	
Subject Sample	# 2	03:49	RFI Detect
Breath Volume	0.000	03:49	
Air Blank	0.000	03:50	
		03:50	RFI Detect
Air Blank	0.000	03:51	

Data Download Date/Time: 04/29/2016 22:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

INSTRUMENT REGISTERED TO: EUSTIS P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWELL,SHAWNA,M
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: HOWELL,SHAWNA
ARRESTING OFFICER AGENCY: Eustis PD
LAST AGENCY INSPECTION: 03/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1787697
DRY GAS EXPIRY DATE: 01/08/2018
VIOLATION CODE: DUI
ARREST TIME: 02:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:13

SUBJECT'S LAST NAME: BAEZ

DATE/TIME OF TEST: 03/18/2016 03:58:47

Diagnostics Check	OK	03:55	
Air Blank	0.000	03:56	
Control Test	0.079	03:56	
Air Blank	0.000	03:57	
Subject Sample	# 1	03:57	Subject Test Refused
Breath Volume	0.000	03:57	
Air Blank	0.000	03:57	
Control Test	0.079	03:58	
Air Blank	0.000	03:58	
Diagnostics Check	OK	03:58	

Data Download Date/Time: 04/29/2016 22:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PUF AHL,G,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: ADAMS,J
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 02/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:07

SUBJECT'S LAST NAME: POTTS II

DATE/TIME OF TEST: 03/10/2016 01:34:28

Diagnostics Check	OK	01:28
Air Blank	0.000	01:28
Control Test	0.079	01:28
Air Blank	0.000	01:29
Subject Sample # 1	0.193	01:29
Breath Volume	2.898	01:29
Air Blank	0.000	01:30
Air Blank	0.000	01:32
Subject Sample # 2	0.186	01:32
Breath Volume	1.832	01:32
Air Blank	0.000	01:33
Control Test	0.079	01:33
Air Blank	0.000	01:34
Diagnostics Check	OK	01:34

Data Download Date/Time: 03/25/2016 11:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,JIMMY,J
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: ADAMS,JIMMY
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 02/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: WEST

DATE/TIME OF TEST: 03/18/2016 00:26:33

Diagnostics Check	OK	00:20
Air Blank	0.000	00:20
Control Test	0.079	00:21
Air Blank	0.000	00:21
Subject Sample # 1	0.181	00:22
Breath Volume	3.535	00:22
Air Blank	0.000	00:22
Air Blank	0.000	00:24
Subject Sample # 2	0.179	00:25
Breath Volume	3.695	00:25
Air Blank	0.000	00:25
Control Test	0.078	00:25
Air Blank	0.000	00:26
Diagnostics Check	OK	00:26

Data Download Date/Time: 03/25/2016 11:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KERN,DEPUTY,B
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: ADAMS,DEPUTY
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 03/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: 0101152017278088=230

DATE/TIME OF TEST: 03/26/2016 02:26:33

Diagnostics Check		OK	02:20
Air Blank		0.000	02:20
Control Test		0.079	02:21
Air Blank		0.000	02:21
Subject Sample	# 1	0.271	02:22
Breath Volume		2.210	02:22
Air Blank		0.000	02:22
Air Blank		0.000	02:24
Subject Sample	# 2	0.256	02:25
Breath Volume		1.746	02:25
Air Blank		0.000	02:25
Control Test		0.078	02:25
Air Blank		0.000	02:26
Diagnostics Check		OK	02:26

Data Download Date/Time: 04/28/2016 09:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KERN,DEPUTY,B
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: KOSTYAL,DEPUTY
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 03/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: EVANS

DATE/TIME OF TEST: 03/31/2016 01:06:28

Diagnostics Check	OK	00:59
Air Blank	0.000	01:00
Control Test	0.079	01:00
Air Blank	0.000	01:01
Subject Sample # 1	0.084	01:01
Breath Volume	1.785	01:01
Air Blank	0.000	01:02
Air Blank	0.000	01:04
Subject Sample # 2	0.081	01:04
Breath Volume	2.113	01:04
Air Blank	0.000	01:05
Control Test	0.078	01:05
Air Blank	0.000	01:06
Diagnostics Check	OK	01:06

Data Download Date/Time: 04/28/2016 09:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SOBAS,PATRICIA,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: KOSTYAL,M
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 02/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: OKOLOWITCZ

DATE/TIME OF TEST: 03/05/2016 03:39:32

Diagnostics Check	OK	03:33
Air Blank	0.000	03:33
Control Test	0.080	03:33
Air Blank	0.000	03:34
Subject Sample # 1	0.074	03:34
Breath Volume	3.058	03:34
Air Blank	0.000	03:35
Air Blank	0.000	03:37
Subject Sample # 2	0.070	03:38
Breath Volume	3.023	03:38
Air Blank	0.000	03:38
Control Test	0.078	03:38
Air Blank	0.000	03:39
Diagnostics Check	OK	03:39

Data Download Date/Time: 03/25/2016 09:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PUFAHL,G,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: PUFAHL,G
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 02/23/2016

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:19

SUBJECT'S LAST NAME: ARSLANAGIC

DATE/TIME OF TEST: 03/05/2016 03:53:36

Diagnostics Check	OK	03:46
Air Blank	0.000	03:46
Control Test	0.080	03:47
Air Blank	0.000	03:47
Subject Sample # 1	0.194	03:48
Breath Volume	1.730	03:48
Air Blank	0.000	03:49
Air Blank	0.000	03:51
Subject Sample # 2	0.195	03:52
Breath Volume	1.750	03:52
Air Blank	0.000	03:52
Control Test	0.079	03:53
Air Blank	0.000	03:53
Diagnostics Check	OK	03:53

Data Download Date/Time: 03/25/2016 09:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,DANIEL,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: SLOWICK,OFFICER
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 02/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 18:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:36

SUBJECT'S LAST NAME: OCHANKOWSKI

DATE/TIME OF TEST: 03/05/2016 20:11:51

Diagnostics Check		OK	19:59	
Air Blank		0.000	19:59	
Control Test		0.080	20:00	
Air Blank		0.000	20:00	
Subject Sample	# 1	0.093	20:03	Volume Not Met
Breath Volume		0.507	20:03	
Air Blank		0.000	20:04	
Air Blank		0.000	20:06	
Subject Sample	# 2	0.101	20:06	
Breath Volume		1.273	20:06	
Air Blank		0.000	20:07	
Air Blank		0.000	20:09	
Subject Sample	# 3	0.101	20:10	
Breath Volume		1.402	20:10	
Air Blank		0.000	20:10	
Control Test		0.079	20:11	
Air Blank		0.000	20:11	
Diagnostics Check		OK	20:11	

Data Download Date/Time: 03/25/2016 09:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARTHUR,MELVIN,J
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: ARTHUR,MELVIN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 02/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:32

SUBJECT'S LAST NAME: GRAY

DATE/TIME OF TEST: 03/06/2016 01:04:16

Diagnostics Check		OK	00:52	
Air Blank		0.000	00:53	
Control Test		0.080	00:53	
Air Blank		0.000	00:53	
Subject Sample	# 1	0.192	00:57	Volume Not Met
Breath Volume		0.570	00:57	
Air Blank		0.000	00:57	
Air Blank		0.000	00:59	
Subject Sample	# 2	0.167	01:02	Volume Not Met
Breath Volume		0.230	01:02	
Air Blank		0.000	01:03	
Control Test		0.080	01:03	
Air Blank		0.000	01:04	
Diagnostics Check		OK	01:04	

Data Download Date/Time: 03/25/2016 09:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANIK,JONATHAN,F
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: VANIK,JONATHAN
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 02/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: MAYHEW

DATE/TIME OF TEST: 03/08/2016 03:47:05

Diagnostics Check		OK	03:35	
Air Blank		0.000	03:35	
Control Test		0.081	03:36	
Air Blank		0.000	03:36	
Subject Sample	# 1	0.162	03:39	Volume Not Met
Breath Volume		0.253	03:39	
Air Blank		0.000	03:40	
Air Blank		0.000	03:42	
Subject Sample	# 2	0.176	03:45	Volume Not Met
Breath Volume		0.082	03:45	
Air Blank		0.000	03:46	
Control Test		0.080	03:46	
Air Blank		0.000	03:47	
Diagnostics Check		OK	03:47	

Data Download Date/Time: 03/25/2016 09:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANIK,JONATHAN,F
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: VANIK,JONATHAN
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 02/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: MAYHEW

DATE/TIME OF TEST: 03/08/2016 04:02:16

Diagnostics Check		OK	03:50	
Air Blank		0.000	03:51	
Control Test		0.079	03:51	
Air Blank		0.000	03:52	
Subject Sample	# 1	0.193	03:55	Volume Not Met
Breath Volume		0.058	03:55	
Air Blank		0.000	03:55	
Air Blank		0.000	03:57	
Subject Sample	# 2	0.186	04:00	Volume Not Met
Breath Volume		0.125	04:00	
Air Blank		0.000	04:01	
Control Test		0.078	04:01	
Air Blank		0.000	04:02	
Diagnostics Check		OK	04:02	

Data Download Date/Time: 03/25/2016 09:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NICK,A.,
BREATH TEST OPERATOR AGENCY: North Port PD
ARRESTING OFFICER: NICK,A.
ARRESTING OFFICER AGENCY: North Port PD
LAST AGENCY INSPECTION: 02/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: HOWARTH

DATE/TIME OF TEST: 03/10/2016 00:15:04

Diagnostics Check	OK	00:07
Air Blank	0.000	00:07
Control Test	0.080	00:07
Air Blank	0.000	00:08
Subject Sample # 1	0.190	00:09
Breath Volume	1.656	00:09
Air Blank	0.000	00:09
Air Blank	0.000	00:11
Subject Sample # 2	0.192	00:13
Breath Volume	2.386	00:13
Air Blank	0.000	00:14
Control Test	0.080	00:14
Air Blank	0.000	00:14
Diagnostics Check	OK	00:15

Data Download Date/Time: 03/25/2016 09:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,NICKKI,N
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BELL,D
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 02/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: STEWART

DATE/TIME OF TEST: 03/11/2016 02:18:21

Diagnostics Check		OK	02:08
Air Blank		0.000	02:08
Control Test		0.081	02:08
Air Blank		0.000	02:09
Subject Sample	# 1	0.207	02:10
Breath Volume		2.992	02:10
Air Blank		0.000	02:11
Air Blank		0.000	02:12
Subject Sample	# 2	0.186	02:13
Breath Volume		1.980	02:13
Air Blank		0.000	02:14
Air Blank		0.000	02:16
Subject Sample	# 3	0.205	02:16
Breath Volume		3.156	02:16
Air Blank		0.000	02:17
Control Test		0.080	02:17
Air Blank		0.000	02:18
Diagnostics Check		OK	02:18

Data Download Date/Time: 03/25/2016 09:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PUF AHL,G,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: PUF AHL,G
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 02/23/2016

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: WOOTEN

DATE/TIME OF TEST: 03/12/2016 04:48:58

Diagnostics Check	OK	04:40
Air Blank	0.000	04:41
Control Test	0.081	04:41
Air Blank	0.000	04:42
Subject Sample # 1	0.165	04:43
Breath Volume	3.632	04:43
Air Blank	0.000	04:44
Air Blank	0.000	04:46
Subject Sample # 2	0.153	04:47
Breath Volume	2.062	04:47
Air Blank	0.000	04:48
Control Test	0.079	04:48
Air Blank	0.000	04:48
Diagnostics Check	OK	04:48

Data Download Date/Time: 03/25/2016 09:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MINGS,DEPUTY G.,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: MINGS,DEPUTY G.
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 02/23/2016

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:11

SUBJECT'S LAST NAME: ANGELASTRI

DATE/TIME OF TEST: 03/14/2016 22:40:56

Diagnostics Check	OK	22:33
Air Blank	0.000	22:34
Control Test	0.080	22:34
Air Blank	0.000	22:34
Subject Sample # 1	0.176	22:36
Breath Volume	1.625	22:36
Air Blank	0.000	22:36
Air Blank	0.000	22:38
Subject Sample # 2	0.176	22:39
Breath Volume	1.750	22:39
Air Blank	0.000	22:40
Control Test	0.080	22:40
Air Blank	0.000	22:40
Diagnostics Check	OK	22:40

Data Download Date/Time: 03/25/2016 09:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KENNEDY,DAVID,A
BREATH TEST OPERATOR AGENCY: Sarasota PD
ARRESTING OFFICER: PULLARA,ANTHONY
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 02/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:30

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 03/16/2016 07:01:16

Diagnostics Check	OK	06:54
Air Blank	0.000	06:55
Control Test	0.079	06:55
Air Blank	0.000	06:55
Subject Sample # 1	0.132	06:56
Breath Volume	2.121	06:56
Air Blank	0.000	06:57
Air Blank	0.000	06:59
Subject Sample # 2	0.130	06:59
Breath Volume	1.789	06:59
Air Blank	0.000	07:00
Control Test	0.079	07:00
Air Blank	0.000	07:01
Diagnostics Check	OK	07:01

Data Download Date/Time: 03/25/2016 09:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANIK,JONATHAN,F
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: VANIK,JONATHAN
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 02/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:57

SUBJECT'S LAST NAME: KLUCITAS

DATE/TIME OF TEST: 03/20/2016 02:31:50

Diagnostics Check		OK	02:20	
Air Blank		0.000	02:20	
Control Test		0.080	02:20	
Air Blank		0.000	02:21	
Subject Sample	# 1	0.204	02:24	Volume Not Met
Breath Volume		0.382	02:24	
Air Blank		0.000	02:25	
Air Blank		0.000	02:27	
Subject Sample	# 2	0.199	02:30	Volume Not Met
Breath Volume		0.683	02:30	
Air Blank		0.000	02:30	
Control Test		0.076	02:31	
Air Blank		0.000	02:31	
Diagnostics Check		OK	02:31	

Data Download Date/Time: 03/25/2016 09:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANIK,JONATHAN,F
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: VANIK,JONATHAN
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 02/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:56

SUBJECT'S LAST NAME: KLUCITAS

DATE/TIME OF TEST: 03/20/2016 02:44:03

Diagnostics Check	OK	02:33	
Air Blank	0.000	02:34	
Control Test	0.075	02:34	
Air Blank	0.000	02:34	
Subject Sample	# 1 0.200	02:38	Volume Not Met
Breath Volume	0.574	02:38	
Air Blank	0.000	02:38	
Air Blank	0.000	02:40	
Subject Sample	# 2	02:42	Subject Test Refused
Breath Volume	0.058	02:42	
Air Blank	0.000	02:43	
Control Test	0.075	02:43	
Air Blank	0.000	02:43	
Diagnostics Check	OK	02:44	

Data Download Date/Time: 03/25/2016 09:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEAVER,RICHARD,A
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: WEAVER,RICHARD
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 02/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 16:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 03/20/2016 20:27:10

Diagnostics Check	OK	20:21
Air Blank	0.000	20:21
Control Test	0.080	20:21
Air Blank	0.000	20:22
Subject Sample # 1	0.000	20:22
Breath Volume	2.707	20:22
Air Blank	0.000	20:23
Air Blank	0.000	20:25
Subject Sample # 2	0.000	20:25
Breath Volume	2.218	20:25
Air Blank	0.000	20:26
Control Test	0.079	20:26
Air Blank	0.000	20:27
Diagnostics Check	OK	20:27

Data Download Date/Time: 03/25/2016 09:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,DANIEL,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: MORALES,OFFICER
ARRESTING OFFICER AGENCY: Venice PD
LAST AGENCY INSPECTION: 02/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:32

SUBJECT'S LAST NAME: THOPSEY

DATE/TIME OF TEST: 03/21/2016 01:28:46

Diagnostics Check	OK	01:23	
Air Blank	0.000	01:24	
Control Test	0.080	01:24	
Air Blank	0.000	01:25	
Subject Sample	# 1	01:27	Subject Test Refused
Breath Volume	0.000	01:27	
Air Blank	0.000	01:27	
Control Test	0.081	01:28	
Air Blank	0.000	01:28	
Diagnostics Check	OK	01:28	

Data Download Date/Time: 03/25/2016 09:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001350

INSTRUMENT REGISTERED TO: FRUITLAND PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAINS,HENRY,E
BREATH TEST OPERATOR AGENCY: Fruitland Park PD
ARRESTING OFFICER: JOHNSON,BARNEY
ARRESTING OFFICER AGENCY: Lady Lake PD
LAST AGENCY INSPECTION: 03/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A1
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: OVERACKER

DATE/TIME OF TEST: 03/18/2016 03:43:24

Diagnostics Check	OK	03:37
Air Blank	0.000	03:37
Control Test	0.080	03:37
Air Blank	0.000	03:38
Subject Sample # 1	0.000	03:39
Breath Volume	2.425	03:39
Air Blank	0.000	03:39
Air Blank	0.000	03:41
Subject Sample # 2	0.000	03:42
Breath Volume	2.554	03:42
Air Blank	0.000	03:42
Control Test	0.078	03:42
Air Blank	0.000	03:43
Diagnostics Check	OK	03:43

Data Download Date/Time: 04/11/2016 15:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001365

INSTRUMENT REGISTERED TO: LEVY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAYFIELD,TOM,F
BREATH TEST OPERATOR AGENCY: Levy County SO
ARRESTING OFFICER: WEATHERFORD,K
ARRESTING OFFICER AGENCY: Levy County SO
LAST AGENCY INSPECTION: 02/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG425202
DRY GAS EXPIRY DATE: 09/09/2016
VIOLATION CODE: DUI
ARREST TIME: 19:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:15

SUBJECT'S LAST NAME: FINE

DATE/TIME OF TEST: 03/12/2016 20:51:02

Diagnostics Check	OK	20:39	
Air Blank	0.000	20:40	
Control Test	0.075	20:40	
Air Blank	0.000	20:41	
Subject Sample	# 1	20:44	No Sample Provided
Breath Volume	0.000	20:44	
Air Blank	0.000	20:44	
Air Blank	0.000	20:46	
Subject Sample	# 2	20:49	Subject Test Refused
Breath Volume	0.000	20:49	
Air Blank	0.000	20:50	
Control Test	0.075	20:50	
Air Blank	0.000	20:50	
Diagnostics Check	OK	20:51	

Data Download Date/Time: 03/29/2016 18:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001365

INSTRUMENT REGISTERED TO: LEVY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BEAUCHAMP,ROSE,M
BREATH TEST OPERATOR AGENCY: Levy County SO
ARRESTING OFFICER: MCNEIL,MICHAEL
ARRESTING OFFICER AGENCY: Levy County SO
LAST AGENCY INSPECTION: 02/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG425202
DRY GAS EXPIRY DATE: 09/09/2016
VIOLATION CODE: DUI
ARREST TIME: 23:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: ROY

DATE/TIME OF TEST: 03/18/2016 01:18:10

Diagnostics Check	OK	01:15	
Air Blank	0.000	01:15	
Control Test	0.079	01:16	
Air Blank	0.000	01:16	
Subject Sample	# 1	01:16	Subject Test Refused
Breath Volume	0.000	01:16	
Air Blank	0.000	01:17	
Control Test	0.079	01:17	
Air Blank	0.000	01:18	
Diagnostics Check	OK	01:18	

Data Download Date/Time: 03/29/2016 18:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001365

INSTRUMENT REGISTERED TO: LEVY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LESHER, TIFFANY, A
BREATH TEST OPERATOR AGENCY: Levy County SO
ARRESTING OFFICER: COSSEBOOM, MIKE
ARRESTING OFFICER AGENCY: Chiefland PD
LAST AGENCY INSPECTION: 02/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG425202
DRY GAS EXPIRY DATE: 09/09/2016
VIOLATION CODE: DUI
ARREST TIME: 21:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 03/24/2016 22:22:22

Diagnostics Check	OK	22:19	
Air Blank	0.000	22:19	
Control Test	0.080	22:20	
Air Blank	0.000	22:20	
Subject Sample	# 1	22:20	Subject Test Refused
Breath Volume	0.000	22:20	
Air Blank	0.000	22:21	
Control Test	0.080	22:21	
Air Blank	0.000	22:22	
Diagnostics Check	OK	22:22	

Data Download Date/Time: 03/29/2016 18:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLE,S,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HOLE,S
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:18

SUBJECT'S LAST NAME: PURPURA

DATE/TIME OF TEST: 03/04/2016 01:53:44

Diagnostics Check	OK	01:47
Air Blank	0.000	01:48
Control Test	0.077	01:48
Air Blank	0.000	01:48
Subject Sample # 1	0.132	01:49
Breath Volume	2.410	01:49
Air Blank	0.000	01:49
Air Blank	0.000	01:51
Subject Sample # 2	0.119	01:52
Breath Volume	1.234	01:52
Air Blank	0.000	01:52
Control Test	0.077	01:53
Air Blank	0.000	01:53
Diagnostics Check	OK	01:53

Data Download Date/Time: 03/25/2016 23:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALATI,VINCENZO,F
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GALATI,VINCENZO
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 02:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:26

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 03/05/2016 03:00:23

Diagnostics Check	OK	02:51	
Air Blank	0.000	02:52	
Control Test	0.079	02:52	
Air Blank	0.000	02:52	
Subject Sample	# 1	02:56	No Sample Provided
Breath Volume	0.000	02:56	
Air Blank	0.000	02:56	
Air Blank	0.000	02:58	
Subject Sample	# 2	02:59	Subject Test Refused
Breath Volume	0.000	02:59	
Air Blank	0.000	02:59	
Control Test	0.078	02:59	
Air Blank	0.000	03:00	
Diagnostics Check	OK	03:00	

Data Download Date/Time: 03/25/2016 23:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALATI,VINCENZO,F
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GALATI,VINCENZO
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 02:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:44

SUBJECT'S LAST NAME: RODRIGUEZ RODRIGUEZ

DATE/TIME OF TEST: 03/06/2016 03:18:58

Diagnostics Check	OK	03:09
Air Blank	0.000	03:09
Control Test	0.078	03:10
Air Blank	0.000	03:10
Subject Sample # 1	0.186	03:13
Breath Volume	1.699	03:13
Air Blank	0.000	03:14
Air Blank	0.000	03:16
Subject Sample # 2	0.183	03:17
Breath Volume	1.554	03:17
Air Blank	0.000	03:18
Control Test	0.078	03:18
Air Blank	0.000	03:18
Diagnostics Check	OK	03:18

Data Download Date/Time: 03/25/2016 23:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LACHANCE,SERGEANT,D
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GEOGHEGAN,DEPUTY
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 11:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:25

SUBJECT'S LAST NAME: FAUGHT

DATE/TIME OF TEST: 03/07/2016 12:46:38

Diagnostics Check	OK	12:46	
Air Blank		12:46	RFI Detect
Air Blank	0.000	12:46	

Data Download Date/Time: 03/25/2016 23:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LACHANCE,SERGEANT,D
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GEOGHEGAN,DEPUTY
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 11:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:25

SUBJECT'S LAST NAME: FAUGHT

DATE/TIME OF TEST: 03/07/2016 12:54:17

Diagnostics Check	OK	12:48
Air Blank	0.000	12:48
Control Test	0.078	12:48
Air Blank	0.000	12:49
Subject Sample # 1	0.161	12:49
Breath Volume	1.628	12:49
Air Blank	0.000	12:50
Air Blank	0.000	12:52
Subject Sample # 2	0.160	12:52
Breath Volume	1.660	12:52
Air Blank	0.000	12:53
Control Test	0.078	12:53
Air Blank	0.000	12:54
Diagnostics Check	OK	12:54

Data Download Date/Time: 03/25/2016 23:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LACHANCE,SERGEANT,D
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GRUNDMANN,DEPUTY
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 12:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:06

SUBJECT'S LAST NAME: OBECK

DATE/TIME OF TEST: 03/07/2016 13:37:00

Diagnostics Check	OK	13:26	
Air Blank	0.000	13:26	
Control Test	0.078	13:27	
Air Blank	0.000	13:27	
Subject Sample	# 1 0.213	13:30	Volume Not Met
Breath Volume	0.105	13:30	
Air Blank	0.000	13:31	
Air Blank	0.000	13:33	
Subject Sample	# 2	13:35	Subject Test Refused
Breath Volume	0.015	13:35	
Air Blank	0.000	13:36	
Control Test	0.077	13:36	
Air Blank	0.000	13:36	
Diagnostics Check	OK	13:37	

Data Download Date/Time: 03/25/2016 23:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAIR, RONALD, J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAIR, RONALD
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:36

SUBJECT'S LAST NAME: TEMPLE

DATE/TIME OF TEST: 03/11/2016 03:03:06

Diagnostics Check	OK	02:57
Air Blank	0.000	02:57
Control Test	0.078	02:57
Air Blank	0.000	02:58
Subject Sample # 1	0.000	02:58
Breath Volume	2.878	02:58
Air Blank	0.000	02:59
Air Blank	0.000	03:01
Subject Sample # 2	0.000	03:01
Breath Volume	2.714	03:01
Air Blank	0.000	03:02
Control Test	0.079	03:02
Air Blank	0.000	03:03
Diagnostics Check	OK	03:03

Data Download Date/Time: 03/25/2016 23:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALATI,VINCENZO,F
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GALATI,VINCENZO
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 23:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:39

SUBJECT'S LAST NAME: MOYA

DATE/TIME OF TEST: 03/12/2016 01:03:41

Diagnostics Check	OK	01:00	
Air Blank	0.000	01:01	
Control Test	0.078	01:01	
Air Blank	0.000	01:02	
Subject Sample	# 1	01:02	Subject Test Refused
Breath Volume	0.000	01:02	
Air Blank	0.000	01:02	
Control Test	0.078	01:03	
Air Blank	0.000	01:03	
Diagnostics Check	OK	01:03	

Data Download Date/Time: 03/25/2016 23:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALATI,VINCENZO,F
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GALATI,VINCENZO
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 03:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: COOK

DATE/TIME OF TEST: 03/12/2016 04:11:49

Diagnostics Check	OK	04:05
Air Blank	0.000	04:06
Control Test	0.078	04:06
Air Blank	0.000	04:06
Subject Sample # 1	0.121	04:07
Breath Volume	2.792	04:07
Air Blank	0.000	04:07
Air Blank	0.000	04:09
Subject Sample # 2	0.116	04:10
Breath Volume	1.617	04:10
Air Blank	0.000	04:10
Control Test	0.078	04:11
Air Blank	0.000	04:11
Diagnostics Check	OK	04:11

Data Download Date/Time: 03/25/2016 23:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCKENZIE,DEPUTY,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MCKENZIE,DEPUTY
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 13:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:40

SUBJECT'S LAST NAME: TERANTINO

DATE/TIME OF TEST: 03/15/2016 14:07:42

Diagnostics Check	OK	14:00
Air Blank	0.000	14:01
Control Test	0.078	14:01
Air Blank	0.000	14:02
Subject Sample # 1	0.000	14:03
Breath Volume	3.105	14:03
Air Blank	0.000	14:03
Air Blank	0.000	14:05
Subject Sample # 2	0.000	14:06
Breath Volume	3.203	14:06
Air Blank	0.000	14:06
Control Test	0.078	14:07
Air Blank	0.000	14:07
Diagnostics Check	OK	14:07

Data Download Date/Time: 03/25/2016 23:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLE,S,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HOLE,S
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 18:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:05

SUBJECT'S LAST NAME: HOLLAND

DATE/TIME OF TEST: 03/15/2016 19:43:18

Diagnostics Check	OK	19:31	
Air Blank	0.000	19:32	
Control Test	0.078	19:32	
Air Blank	0.000	19:33	
Subject Sample	# 1	19:36	No Sample Provided
Breath Volume	0.000	19:36	
Air Blank	0.000	19:36	
Air Blank	0.000	19:38	
Subject Sample	# 2	19:41	No Sample Provided
Breath Volume	0.000	19:41	
Air Blank	0.000	19:42	
Control Test	0.079	19:42	
Air Blank	0.000	19:43	
Diagnostics Check	OK	19:43	

Data Download Date/Time: 03/25/2016 23:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILHELM,JAMES,D
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TRINIDAD,MICHAEL
ARRESTING OFFICER AGENCY: Tarpon Springs PD
LAST AGENCY INSPECTION: 02/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 01:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:08

SUBJECT'S LAST NAME: JESEK

DATE/TIME OF TEST: 03/16/2016 02:36:08

Diagnostics Check	OK	02:31	
Air Blank	0.000	02:31	
Control Test	0.078	02:32	
Air Blank	0.000	02:32	
Subject Sample	# 1	02:34	Subject Test Refused
Breath Volume	0.058	02:34	
Air Blank	0.000	02:35	
Control Test	0.080	02:35	
Air Blank	0.000	02:36	
Diagnostics Check	OK	02:36	

Data Download Date/Time: 03/25/2016 23:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALATI,VINCENZO,F
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GALATI,VINCENZO
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 22:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:32

SUBJECT'S LAST NAME: ROJAS-MARQUEZ

DATE/TIME OF TEST: 03/20/2016 00:04:11

Diagnostics Check	OK	23:57
Air Blank	0.000	23:57
Control Test	0.078	23:58
Air Blank	0.000	23:58
Subject Sample # 2	0.201	23:59
Breath Volume	1.742	23:59
Air Blank	0.000	23:59
Air Blank	0.000	00:01
Subject Sample # 1	0.184	00:02
Breath Volume	1.957	00:02
Air Blank	0.000	00:03
Control Test	0.078	00:03
Air Blank	0.000	00:04
Diagnostics Check	OK	00:04

Data Download Date/Time: 03/25/2016 23:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALATI,VINCENZO,F
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GALATI,VINCENZO
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 03:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: TIEDEMANN

DATE/TIME OF TEST: 03/20/2016 04:18:49

Diagnostics Check	OK	04:07	
Air Blank	0.000	04:07	
Control Test	0.079	04:08	
Air Blank	0.000	04:08	
Subject Sample	# 1	04:11	No Sample Provided
Breath Volume	0.000	04:11	
Air Blank	0.000	04:12	
Air Blank	0.000	04:14	
Subject Sample	# 2	04:17	No Sample Provided
Breath Volume	0.000	04:17	
Air Blank	0.000	04:17	
Control Test	0.078	04:18	
Air Blank	0.000	04:18	
Diagnostics Check	OK	04:18	

Data Download Date/Time: 03/25/2016 23:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLAIRE,J,L
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HOOD,C
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 23:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: MATHERNE

DATE/TIME OF TEST: 03/22/2016 00:04:19

Diagnostics Check	OK	23:55
Air Blank	0.000	23:55
Control Test	0.078	23:56
Air Blank	0.000	23:56
Subject Sample # 2	0.210	23:58
Breath Volume	1.402	23:58
Air Blank	0.000	23:59
Air Blank	0.000	00:01
Subject Sample # 1	0.203	00:02
Breath Volume	1.957	00:02
Air Blank	0.000	00:03
Control Test	0.078	00:03
Air Blank	0.000	00:04
Diagnostics Check	OK	00:04

Data Download Date/Time: 03/25/2016 23:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLAIRE,J,L
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GRIFFIN,S
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 03:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:54

SUBJECT'S LAST NAME: HAYES

DATE/TIME OF TEST: 03/22/2016 04:29:20

Diagnostics Check		OK	04:14	
Air Blank		0.000	04:15	
Control Test		0.079	04:15	
Air Blank		0.000	04:16	
Subject Sample	# 1	0.209	04:18	
Breath Volume		1.312	04:18	
Air Blank		0.000	04:19	
Air Blank		0.000	04:20	
Subject Sample	# 2	0.157	04:24	Volume Not Met
Breath Volume		0.820	04:24	
Air Blank		0.000	04:24	
Air Blank		0.000	04:26	
Subject Sample	# 3	0.192	04:27	
Breath Volume		1.511	04:27	
Air Blank		0.000	04:28	
Control Test		0.077	04:28	
Air Blank		0.000	04:29	
Diagnostics Check		OK	04:29	

Data Download Date/Time: 03/25/2016 23:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALATI,VINCENZO,F
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GALATI,VINCENZO
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 03:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:38

SUBJECT'S LAST NAME: BOTKO

DATE/TIME OF TEST: 03/25/2016 04:04:56

Diagnostics Check		OK	03:58
Air Blank		0.000	03:58
Control Test		0.078	03:58
Air Blank		0.000	03:59
Subject Sample	# 1	0.177	04:00
Breath Volume		1.964	04:00
Air Blank		0.000	04:00
Air Blank		0.000	04:02
Subject Sample	# 2	0.188	04:03
Breath Volume		2.460	04:03
Air Blank		0.000	04:03
Control Test		0.078	04:04
Air Blank		0.000	04:04
Diagnostics Check		OK	04:04

Data Download Date/Time: 03/25/2016 23:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALATI,VINCENZO,F
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GALATI,VINCENZO
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 20:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:11

SUBJECT'S LAST NAME: CROSBY

DATE/TIME OF TEST: 03/25/2016 21:49:36

Diagnostics Check	OK	21:44	
Air Blank	0.000	21:45	
Control Test	0.078	21:45	
Air Blank	0.000	21:46	
Subject Sample	# 1	21:48	Subject Test Refused
Breath Volume	0.000	21:48	
Air Blank	0.000	21:48	
Control Test	0.078	21:48	
Air Blank	0.000	21:49	
Diagnostics Check	OK	21:49	

Data Download Date/Time: 03/25/2016 23:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALATI,VINCENZO,F
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GALATI,VINCENZO
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 04:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:54

SUBJECT'S LAST NAME: RUSSELL JR

DATE/TIME OF TEST: 03/27/2016 05:22:04

Diagnostics Check	OK	05:14
Air Blank	0.000	05:15
Control Test	0.077	05:15
Air Blank	0.000	05:15
Subject Sample # 1	0.185	05:17
Breath Volume	1.730	05:17
Air Blank	0.000	05:18
Air Blank	0.000	05:19
Subject Sample # 2	0.167	05:20
Breath Volume	1.562	05:20
Air Blank	0.000	05:21
Control Test	0.078	05:21
Air Blank	0.000	05:21
Diagnostics Check	OK	05:22

Data Download Date/Time: 04/23/2016 20:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLAIRE,J,L
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WEDE,B
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 12:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: ADLEMAN

DATE/TIME OF TEST: 03/28/2016 01:22:19

Diagnostics Check	OK	01:10	
Air Blank	0.000	01:11	
Control Test	0.079	01:11	
Air Blank	0.000	01:12	
Subject Sample	# 1	01:15	No Sample Provided
Breath Volume	0.000	01:15	
Air Blank	0.000	01:15	
Air Blank	0.000	01:17	
Subject Sample	# 2	01:20	No Sample Provided
Breath Volume	0.000	01:20	
Air Blank	0.000	01:21	
Control Test	0.078	01:21	
Air Blank	0.000	01:22	
Diagnostics Check	OK	01:22	

Data Download Date/Time: 04/23/2016 20:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001368

INSTRUMENT REGISTERED TO: NAS MAYPORT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OCASIO,JOSE,A
BREATH TEST OPERATOR AGENCY: Mayport Naval Base
ARRESTING OFFICER: POLLOCK,DEREK
ARRESTING OFFICER AGENCY: Mayport Naval Base
LAST AGENCY INSPECTION: 02/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A1
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: MOTA

DATE/TIME OF TEST: 03/27/2016 03:27:50

Diagnostics Check	OK	03:20
Air Blank	0.000	03:20
Control Test	0.079	03:21
Air Blank	0.000	03:21
Subject Sample # 1	0.159	03:23
Breath Volume	2.015	03:23
Air Blank	0.000	03:23
Air Blank	0.000	03:25
Subject Sample # 2	0.149	03:26
Breath Volume	1.207	03:26
Air Blank	0.000	03:26
Control Test	0.078	03:27
Air Blank	0.000	03:27
Diagnostics Check	OK	03:27

Data Download Date/Time: 03/31/2016 08:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CANER, OSVALDO,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: WEAVER, WILLIAM
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 02/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:04

SUBJECT'S LAST NAME: NEWHARD

DATE/TIME OF TEST: 03/01/2016 21:32:23

Diagnostics Check	OK	21:26
Air Blank	0.000	21:26
Control Test	0.082	21:26
Air Blank	0.000	21:27
Subject Sample # 1	0.215	21:28
Breath Volume	2.839	21:28
Air Blank	0.000	21:28
Air Blank	0.000	21:30
Subject Sample # 2	0.213	21:30
Breath Volume	2.035	21:30
Air Blank	0.000	21:31
Control Test	0.082	21:31
Air Blank	0.000	21:32
Diagnostics Check	OK	21:32

Data Download Date/Time: 03/03/2016 15:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,COURTNEY,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: LAL,RAJIV
ARRESTING OFFICER AGENCY: Winter Garden PD
LAST AGENCY INSPECTION: 02/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 19:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:09

SUBJECT'S LAST NAME: ZEIGLER

DATE/TIME OF TEST: 03/02/2016 20:49:03

Diagnostics Check	OK	20:37	
Air Blank	0.000	20:38	
Control Test	0.081	20:38	
Air Blank	0.000	20:38	
Subject Sample	# 1	20:42	No Sample Provided
Breath Volume	0.000	20:42	
Air Blank	0.000	20:42	
Air Blank	0.000	20:44	
Subject Sample	# 2	20:47	No Sample Provided
Breath Volume	0.000	20:47	
Air Blank	0.000	20:48	
Control Test	0.082	20:48	
Air Blank	0.000	20:48	
Diagnostics Check	OK	20:49	

Data Download Date/Time: 03/03/2016 15:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCREWS,DARCY,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: PRICE,FORREST
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 02/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: CARTER

DATE/TIME OF TEST: 03/03/2016 03:16:03

Diagnostics Check		OK	03:09
Air Blank		0.000	03:10
Control Test		0.082	03:10
Air Blank		0.000	03:10
Subject Sample	# 1	0.254	03:11
Breath Volume		1.906	03:11
Air Blank		0.000	03:12
Air Blank		0.000	03:14
Subject Sample	# 2	0.266	03:14
Breath Volume		3.300	03:14
Air Blank		0.000	03:15
Control Test		0.081	03:15
Air Blank		0.000	03:15
Diagnostics Check		OK	03:16

Data Download Date/Time: 03/03/2016 15:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITFIELD, TIARA, D
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ANDREWS, GARY
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 02/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: ANDREWS

DATE/TIME OF TEST: 03/04/2016 04:15:42

Diagnostics Check	OK	04:09
Air Blank	0.000	04:09
Control Test	0.082	04:09
Air Blank	0.000	04:10
Subject Sample # 1	0.171	04:10
Breath Volume	1.906	04:10
Air Blank	0.000	04:11
Air Blank	0.000	04:13
Subject Sample # 2	0.174	04:14
Breath Volume	1.898	04:14
Air Blank	0.000	04:14
Control Test	0.080	04:15
Air Blank	0.000	04:15
Diagnostics Check	OK	04:15

Data Download Date/Time: 03/10/2016 16:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,COURTNEY,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: GABRUS,KYLE
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 02/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:41

SUBJECT'S LAST NAME: WEDGE

DATE/TIME OF TEST: 03/05/2016 02:19:20

Diagnostics Check	OK	02:07	
Air Blank	0.000	02:08	
Control Test	0.082	02:08	
Air Blank	0.000	02:09	
Subject Sample	# 1	02:12	No Sample Provided
Breath Volume	0.000	02:12	
Air Blank	0.000	02:12	
Air Blank	0.000	02:14	
Subject Sample	# 2	02:17	No Sample Provided
Breath Volume	0.000	02:17	
Air Blank	0.000	02:18	
Control Test	0.081	02:18	
Air Blank	0.000	02:19	
Diagnostics Check	OK	02:19	

Data Download Date/Time: 03/10/2016 16:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLEMAN,BRETT,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DOBBINS,SEAN
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 02/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: NACHT

DATE/TIME OF TEST: 03/05/2016 03:07:57

Diagnostics Check	OK	03:02
Air Blank	0.000	03:02
Control Test	0.082	03:02
Air Blank	0.000	03:03
Subject Sample # 1	0.172	03:03
Breath Volume	2.242	03:03
Air Blank	0.000	03:04
Air Blank	0.000	03:06
Subject Sample # 2	0.172	03:06
Breath Volume	2.988	03:06
Air Blank	0.000	03:07
Control Test	0.081	03:07
Air Blank	0.000	03:07
Diagnostics Check	OK	03:07

Data Download Date/Time: 03/10/2016 16:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,COURTNEY,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: GONZALEZ,AGUSTIN
ARRESTING OFFICER AGENCY: Central FL Univ PD
LAST AGENCY INSPECTION: 02/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:53

SUBJECT'S LAST NAME: GUISAO PEREZ

DATE/TIME OF TEST: 03/05/2016 04:31:03

Diagnostics Check	OK	04:18	
Air Blank	0.000	04:18	
Control Test	0.082	04:18	
Air Blank	0.000	04:19	
Subject Sample	# 1 0.151	04:21	
Breath Volume	3.437	04:21	
Air Blank	0.000	04:21	
Air Blank	0.000	04:23	
Subject Sample	# 2	04:26	No Sample Provided
Breath Volume	0.000	04:26	
Air Blank	0.000	04:27	
Air Blank	0.000	04:29	
Subject Sample	# 3 0.134	04:29	
Breath Volume	1.804	04:29	
Air Blank	0.000	04:30	
Control Test	0.081	04:30	
Air Blank	0.000	04:30	
Diagnostics Check	OK	04:31	

Data Download Date/Time: 03/10/2016 16:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLEMAN,BRETT,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HAWKINS,DEBORAH
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 02/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:53

SUBJECT'S LAST NAME: GORDIANO

DATE/TIME OF TEST: 03/05/2016 05:23:40

Diagnostics Check	OK	05:17
Air Blank	0.000	05:18
Control Test	0.082	05:18
Air Blank	0.000	05:19
Subject Sample # 1	0.000	05:19
Breath Volume	4.277	05:19
Air Blank	0.000	05:19
Air Blank	0.000	05:21
Subject Sample # 2	0.000	05:22
Breath Volume	4.019	05:22
Air Blank	0.000	05:22
Control Test	0.082	05:23
Air Blank	0.000	05:23
Diagnostics Check	OK	05:23

Data Download Date/Time: 03/10/2016 16:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FIGUEROA, HECTOR, A
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ANDREWS, GARY
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 02/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:38

SUBJECT'S LAST NAME: ALDANA ARIAS

DATE/TIME OF TEST: 03/06/2016 04:10:01

Diagnostics Check		OK	04:00
Air Blank		0.000	04:01
Control Test		0.082	04:01
Air Blank		0.000	04:02
Subject Sample	# 1	0.151	04:02
Breath Volume		1.828	04:02
Air Blank		0.000	04:03
Air Blank		0.000	04:05
Subject Sample	# 2	0.176	04:05
Breath Volume		2.519	04:05
Air Blank		0.000	04:06
Air Blank		0.000	04:08
Subject Sample	# 3	0.157	04:08
Breath Volume		1.308	04:08
Air Blank		0.000	04:09
Control Test		0.081	04:09
Air Blank		0.000	04:09
Diagnostics Check		OK	04:10

Data Download Date/Time: 03/10/2016 16:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLEMAN,BRETT,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: REMBISLICARI,JAZMIN
ARRESTING OFFICER AGENCY: Ocoee PD
LAST AGENCY INSPECTION: 02/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: STEWART

DATE/TIME OF TEST: 03/07/2016 03:28:16

Diagnostics Check		OK	03:13	
Air Blank		0.000	03:14	
Control Test		0.081	03:14	
Air Blank		0.000	03:15	
Subject Sample	# 1	0.193	03:18	Volume Not Met
Breath Volume		0.066	03:18	
Air Blank		0.000	03:18	
Air Blank		0.000	03:20	
Subject Sample	# 2	0.184	03:21	
Breath Volume		1.242	03:21	
Air Blank		0.000	03:21	
Air Blank		0.000	03:23	
Subject Sample	# 3	0.186	03:26	Volume Not Met
Breath Volume		0.980	03:26	
Air Blank		0.000	03:27	
Control Test		0.081	03:27	
Air Blank		0.000	03:28	
Diagnostics Check		OK	03:28	

Data Download Date/Time: 03/10/2016 16:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,SHAWN,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CONWAY,COREY
ARRESTING OFFICER AGENCY: Winter Park PD
LAST AGENCY INSPECTION: 02/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:52

SUBJECT'S LAST NAME: BAKER

DATE/TIME OF TEST: 03/11/2016 03:22:24

Diagnostics Check	OK	03:15
Air Blank	0.000	03:16
Control Test	0.080	03:16
Air Blank	0.000	03:16
Subject Sample # 1	0.303	03:17
Breath Volume	1.425	03:17
Air Blank	0.000	03:18
Air Blank	0.000	03:20
Subject Sample # 2	0.323	03:20
Breath Volume	2.148	03:20
Air Blank	0.000	03:21
Control Test	0.080	03:21
Air Blank	0.000	03:22
Diagnostics Check	OK	03:22

Data Download Date/Time: 03/17/2016 16:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDEN,WILLIAM,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HAWKINS,DEBBIE
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 02/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:13

SUBJECT'S LAST NAME: HERMAN

DATE/TIME OF TEST: 03/11/2016 03:48:23

Diagnostics Check	OK	03:41
Air Blank	0.000	03:41
Control Test	0.081	03:41
Air Blank	0.000	03:42
Subject Sample # 1	0.193	03:43
Breath Volume	2.570	03:43
Air Blank	0.000	03:44
Air Blank	0.000	03:45
Subject Sample # 2	0.189	03:46
Breath Volume	1.859	03:46
Air Blank	0.000	03:47
Control Test	0.081	03:47
Air Blank	0.000	03:48
Diagnostics Check	OK	03:48

Data Download Date/Time: 03/17/2016 16:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,COURTNEY,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CORSO,ANTHONY
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 02/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 18:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:15

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 03/11/2016 20:45:13

Diagnostics Check	OK	20:39
Air Blank	0.000	20:39
Control Test	0.081	20:39
Air Blank	0.000	20:40
Subject Sample # 1	0.163	20:40
Breath Volume	2.835	20:40
Air Blank	0.000	20:41
Air Blank	0.000	20:43
Subject Sample # 2	0.166	20:43
Breath Volume	2.441	20:43
Air Blank	0.000	20:44
Control Test	0.081	20:44
Air Blank	0.000	20:45
Diagnostics Check	OK	20:45

Data Download Date/Time: 03/17/2016 16:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLEMAN,BRETT,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DOBBINS,SEAN
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 02/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: ORLANDO

DATE/TIME OF TEST: 03/12/2016 00:17:04

Diagnostics Check	OK	00:10
Air Blank	0.000	00:11
Control Test	0.082	00:11
Air Blank	0.000	00:12
Subject Sample # 1	0.118	00:12
Breath Volume	1.746	00:12
Air Blank	0.000	00:13
Air Blank	0.000	00:15
Subject Sample # 2	0.118	00:15
Breath Volume	3.148	00:15
Air Blank	0.000	00:16
Control Test	0.081	00:16
Air Blank	0.000	00:16
Diagnostics Check	OK	00:17

Data Download Date/Time: 03/17/2016 16:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,COURTNEY,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ANDREWS,GARY
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 02/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:07

SUBJECT'S LAST NAME: GRADY

DATE/TIME OF TEST: 03/12/2016 00:37:59

Diagnostics Check		OK	00:31
Air Blank		0.000	00:31
Control Test		0.081	00:31
Air Blank		0.000	00:32
Subject Sample	# 1	0.230	00:33
Breath Volume		2.222	00:33
Air Blank		0.000	00:33
Air Blank		0.000	00:35
Subject Sample	# 2	0.218	00:36
Breath Volume		1.527	00:36
Air Blank		0.000	00:37
Control Test		0.080	00:37
Air Blank		0.000	00:37
Diagnostics Check		OK	00:37

Data Download Date/Time: 03/17/2016 16:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ALEJANDRO, GENESIS
ARRESTING OFFICER AGENCY: Winter Park PD
LAST AGENCY INSPECTION: 02/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:01

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 03/12/2016 01:36:46

Diagnostics Check		OK	01:21	
Air Blank		0.000	01:22	
Control Test		0.081	01:22	
Air Blank		0.000	01:23	
Subject Sample	# 1	0.259	01:25	
Breath Volume		1.750	01:25	
Air Blank		0.000	01:26	
Air Blank		0.000	01:28	
Subject Sample	# 2	0.253	01:31	Volume Not Met
Breath Volume		0.000	01:31	
Air Blank		0.000	01:32	
Air Blank		0.000	01:34	
Subject Sample	# 3	0.264	01:35	
Breath Volume		2.140	01:35	
Air Blank		0.000	01:35	
Control Test		0.081	01:36	
Air Blank		0.000	01:36	
Diagnostics Check		OK	01:36	

Data Download Date/Time: 03/17/2016 16:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLEMAN,BRETT,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: LEWIS,BEN
ARRESTING OFFICER AGENCY: Central FL Univ PD
LAST AGENCY INSPECTION: 02/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:31

SUBJECT'S LAST NAME: ALVAREZPEREZ

DATE/TIME OF TEST: 03/12/2016 02:06:47

Diagnostics Check	OK	01:58
Air Blank	0.000	01:59
Control Test	0.082	01:59
Air Blank	0.000	01:59
Subject Sample # 1	0.275	02:01
Breath Volume	1.449	02:01
Air Blank	0.000	02:01
Air Blank	0.000	02:03
Subject Sample # 2	0.262	02:05
Breath Volume	1.769	02:05
Air Blank	0.000	02:05
Control Test	0.081	02:06
Air Blank	0.000	02:06
Diagnostics Check	OK	02:06

Data Download Date/Time: 03/17/2016 16:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,COURTNEY,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CHAPLIN,CHRIS
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 02/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:26

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 03/12/2016 02:57:42

Diagnostics Check	OK	02:51
Air Blank	0.000	02:51
Control Test	0.082	02:52
Air Blank	0.000	02:52
Subject Sample # 1	0.140	02:53
Breath Volume	3.699	02:53
Air Blank	0.000	02:53
Air Blank	0.000	02:55
Subject Sample # 2	0.139	02:56
Breath Volume	3.683	02:56
Air Blank	0.000	02:56
Control Test	0.081	02:57
Air Blank	0.000	02:57
Diagnostics Check	OK	02:57

Data Download Date/Time: 03/17/2016 16:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: GUY, STEVEN
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 02/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:22

SUBJECT'S LAST NAME: BSHARA

DATE/TIME OF TEST: 03/12/2016 03:53:35

Diagnostics Check	OK	03:47
Air Blank	0.000	03:48
Control Test	0.083	03:48
Air Blank	0.000	03:48
Subject Sample # 1	0.147	03:49
Breath Volume	1.347	03:49
Air Blank	0.000	03:49
Air Blank	0.000	03:51
Subject Sample # 2	0.149	03:52
Breath Volume	1.527	03:52
Air Blank	0.000	03:52
Control Test	0.082	03:53
Air Blank	0.000	03:53
Diagnostics Check	OK	03:53

Data Download Date/Time: 03/17/2016 16:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLEMAN,BRETT,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: IMPARATO,FRANK
ARRESTING OFFICER AGENCY: Central FL Univ PD
LAST AGENCY INSPECTION: 02/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:49

SUBJECT'S LAST NAME: ABRAHAM

DATE/TIME OF TEST: 03/12/2016 04:16:46

Diagnostics Check	OK	04:10
Air Blank	0.000	04:11
Control Test	0.082	04:11
Air Blank	0.000	04:12
Subject Sample # 1	0.000	04:12
Breath Volume	4.234	04:12
Air Blank	0.000	04:12
Air Blank	0.000	04:15
Subject Sample # 2	0.000	04:15
Breath Volume	4.664	04:15
Air Blank	0.000	04:15
Control Test	0.081	04:16
Air Blank	0.000	04:16
Diagnostics Check	OK	04:16

Data Download Date/Time: 03/17/2016 16:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: IV, ANDREW
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 02/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:04

SUBJECT'S LAST NAME: GUZMAN BENITEZ

DATE/TIME OF TEST: 03/12/2016 04:31:11

Diagnostics Check	OK	04:24
Air Blank	0.000	04:25
Control Test	0.081	04:25
Air Blank	0.000	04:26
Subject Sample # 1	0.208	04:26
Breath Volume	1.152	04:26
Air Blank	0.000	04:27
Air Blank	0.000	04:29
Subject Sample # 2	0.223	04:29
Breath Volume	2.195	04:29
Air Blank	0.000	04:30
Control Test	0.081	04:30
Air Blank	0.000	04:31
Diagnostics Check	OK	04:31

Data Download Date/Time: 03/17/2016 16:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLEMAN,BRETT,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: KNIGHT,RYAN
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 02/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: SCHUBERT

DATE/TIME OF TEST: 03/12/2016 04:42:44

Diagnostics Check	OK	04:36
Air Blank	0.000	04:37
Control Test	0.082	04:37
Air Blank	0.000	04:37
Subject Sample # 1	0.224	04:38
Breath Volume	1.910	04:38
Air Blank	0.000	04:38
Air Blank	0.000	04:40
Subject Sample # 2	0.227	04:41
Breath Volume	2.617	04:41
Air Blank	0.000	04:41
Control Test	0.081	04:42
Air Blank	0.000	04:42
Diagnostics Check	OK	04:42

Data Download Date/Time: 03/17/2016 16:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLEMAN,BRETT,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HAWKINS,DEBORAH
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 02/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:48

SUBJECT'S LAST NAME: MINAYA

DATE/TIME OF TEST: 03/12/2016 05:24:56

Diagnostics Check	OK	05:10	
Air Blank	0.000	05:10	
Control Test	0.081	05:10	
Air Blank	0.000	05:11	
Subject Sample # 1	0.221	05:12	
Breath Volume	1.531	05:12	
Air Blank	0.000	05:12	
Air Blank	0.000	05:14	
Subject Sample # 2		05:17	No Sample Provided
Breath Volume	0.000	05:17	
Air Blank	0.000	05:18	
Air Blank	0.000	05:20	
Subject Sample # 3		05:23	No Sample Provided
Breath Volume	0.000	05:23	
Air Blank	0.000	05:24	
Control Test	0.081	05:24	
Air Blank	0.000	05:24	
Diagnostics Check	OK	05:24	

Data Download Date/Time: 03/17/2016 16:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,COURTNEY,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HOLLEY,WILLIAM
ARRESTING OFFICER AGENCY: Winter Park PD
LAST AGENCY INSPECTION: 02/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 18:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:43

SUBJECT'S LAST NAME: BEAUX

DATE/TIME OF TEST: 03/12/2016 20:12:12

Diagnostics Check	OK	20:05
Air Blank	0.000	20:05
Control Test	0.082	20:06
Air Blank	0.000	20:06
Subject Sample # 1	0.191	20:07
Breath Volume	1.179	20:07
Air Blank	0.000	20:08
Air Blank	0.000	20:10
Subject Sample # 2	0.211	20:10
Breath Volume	2.914	20:10
Air Blank	0.000	20:11
Control Test	0.081	20:11
Air Blank	0.000	20:12
Diagnostics Check	OK	20:12

Data Download Date/Time: 03/17/2016 16:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FIGUEROA, HECTOR, A
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: TORRES, GEAN
ARRESTING OFFICER AGENCY: Winter Garden PD
LAST AGENCY INSPECTION: 02/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:17

SUBJECT'S LAST NAME: KELLOGG

DATE/TIME OF TEST: 03/13/2016 05:46:23

Diagnostics Check	OK	05:40
Air Blank	0.000	05:40
Control Test	0.082	05:40
Air Blank	0.000	05:41
Subject Sample # 1	0.241	05:41
Breath Volume	1.632	05:41
Air Blank	0.000	05:42
Air Blank	0.000	05:44
Subject Sample # 2	0.260	05:44
Breath Volume	2.824	05:44
Air Blank	0.000	05:45
Control Test	0.082	05:45
Air Blank	0.000	05:46
Diagnostics Check	OK	05:46

Data Download Date/Time: 03/17/2016 16:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: QUINN,RYAN
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 02/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 15:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:07

SUBJECT'S LAST NAME: HUSIC

DATE/TIME OF TEST: 03/13/2016 16:36:28

Diagnostics Check	OK	16:30
Air Blank	0.000	16:30
Control Test	0.081	16:30
Air Blank	0.000	16:31
Subject Sample # 1	0.224	16:31
Breath Volume	1.445	16:31
Air Blank	0.000	16:32
Air Blank	0.000	16:34
Subject Sample # 2	0.243	16:34
Breath Volume	3.187	16:34
Air Blank	0.000	16:35
Control Test	0.081	16:35
Air Blank	0.000	16:36
Diagnostics Check	OK	16:36

Data Download Date/Time: 03/17/2016 16:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ, MICHAEL, J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: STANALAND, BEN
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 02/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 13:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:20

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 03/14/2016 13:52:39

Diagnostics Check	OK	13:46
Air Blank	0.000	13:46
Control Test	0.081	13:47
Air Blank	0.000	13:47
Subject Sample # 1	0.327	13:48
Breath Volume	1.632	13:48
Air Blank	0.000	13:48
Air Blank	0.000	13:50
Subject Sample # 2	0.317	13:51
Breath Volume	1.941	13:51
Air Blank	0.000	13:51
Control Test	0.082	13:52
Air Blank	0.000	13:52
Diagnostics Check	OK	13:52

Data Download Date/Time: 03/17/2016 16:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ, MICHAEL, J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: JAMES, JOHN
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 02/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 15:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:30

SUBJECT'S LAST NAME: MARTINS

DATE/TIME OF TEST: 03/14/2016 17:00:59

Diagnostics Check	OK	16:54
Air Blank	0.000	16:55
Control Test	0.082	16:55
Air Blank	0.000	16:56
Subject Sample # 1	0.000	16:56
Breath Volume	3.417	16:56
Air Blank	0.000	16:57
Air Blank	0.000	16:59
Subject Sample # 2	0.000	16:59
Breath Volume	2.308	16:59
Air Blank	0.000	17:00
Control Test	0.081	17:00
Air Blank	0.000	17:00
Diagnostics Check	OK	17:00

Data Download Date/Time: 03/17/2016 16:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MELVILLE,KELLY,M
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: WHITMAN,JIM
ARRESTING OFFICER AGENCY: Winter Park PD
LAST AGENCY INSPECTION: 02/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 08:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:21

SUBJECT'S LAST NAME: RUST

DATE/TIME OF TEST: 03/16/2016 10:02:05

Diagnostics Check	OK	09:55
Air Blank	0.000	09:56
Control Test	0.082	09:56
Air Blank	0.000	09:56
Subject Sample # 1	0.000	09:57
Breath Volume	2.507	09:57
Air Blank	0.000	09:58
Air Blank	0.000	10:00
Subject Sample # 2	0.000	10:00
Breath Volume	1.640	10:00
Air Blank	0.000	10:01
Control Test	0.081	10:01
Air Blank	0.000	10:02
Diagnostics Check	OK	10:02

Data Download Date/Time: 03/17/2016 16:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCREWS,DARCY,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CORSO,ANTHONY
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 02/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:12

SUBJECT'S LAST NAME: JOUSMA

DATE/TIME OF TEST: 03/16/2016 23:38:41

Diagnostics Check	OK	23:32
Air Blank	0.000	23:33
Control Test	0.081	23:33
Air Blank	0.000	23:33
Subject Sample # 1	0.127	23:34
Breath Volume	2.109	23:34
Air Blank	0.000	23:34
Air Blank	0.000	23:36
Subject Sample # 2	0.122	23:37
Breath Volume	1.445	23:37
Air Blank	0.000	23:37
Control Test	0.081	23:38
Air Blank	0.000	23:38
Diagnostics Check	OK	23:38

Data Download Date/Time: 03/17/2016 16:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,SHAWN,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: EVANS,JOSH
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 02/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:54

SUBJECT'S LAST NAME: GUILLUAME

DATE/TIME OF TEST: 03/17/2016 04:29:50

Diagnostics Check		OK	04:17	
Air Blank		0.000	04:18	
Control Test		0.082	04:18	
Air Blank		0.000	04:19	
Subject Sample	# 1	0.108	04:22	Volume Not Met
Breath Volume		0.164	04:22	
Air Blank		0.000	04:22	
Air Blank		0.000	04:24	
Subject Sample	# 2	0.125	04:25	
Breath Volume		2.449	04:25	
Air Blank		0.000	04:26	
Air Blank		0.000	04:27	
Subject Sample	# 3	0.126	04:28	
Breath Volume		1.281	04:28	
Air Blank		0.000	04:28	
Control Test		0.081	04:29	
Air Blank		0.000	04:29	
Diagnostics Check		OK	04:29	

Data Download Date/Time: 03/17/2016 16:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FIGUEROA, HECTOR, A
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: GALLUP, ELIAS
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 02/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:28

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 03/18/2016 01:57:39

Diagnostics Check	OK	01:51
Air Blank	0.000	01:52
Control Test	0.082	01:52
Air Blank	0.000	01:52
Subject Sample # 1	0.119	01:53
Breath Volume	1.480	01:53
Air Blank	0.000	01:53
Air Blank	0.000	01:55
Subject Sample # 2	0.121	01:56
Breath Volume	1.390	01:56
Air Blank	0.000	01:56
Control Test	0.081	01:57
Air Blank	0.000	01:57
Diagnostics Check	OK	01:57

Data Download Date/Time: 03/22/2016 11:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,SHAWN,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: IRELAND,DAVID
ARRESTING OFFICER AGENCY: Edgewood PD
LAST AGENCY INSPECTION: 02/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:34

SUBJECT'S LAST NAME: LOCASCIO

DATE/TIME OF TEST: 03/18/2016 03:02:56

Diagnostics Check		OK	02:57
Air Blank		0.000	02:57
Control Test		0.081	02:57
Air Blank		0.000	02:58
Subject Sample	# 1	0.207	02:58
Breath Volume		1.496	02:58
Air Blank		0.000	02:59
Air Blank		0.000	03:01
Subject Sample	# 2	0.211	03:01
Breath Volume		1.683	03:01
Air Blank		0.000	03:02
Control Test		0.081	03:02
Air Blank		0.000	03:02
Diagnostics Check		OK	03:02

Data Download Date/Time: 03/22/2016 11:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FIGUEROA, HECTOR, A
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: EVANS, JOSH
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 02/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: ROOZITALAB

DATE/TIME OF TEST: 03/18/2016 03:29:24

Diagnostics Check		OK	03:22
Air Blank		0.000	03:22
Control Test		0.081	03:23
Air Blank		0.000	03:23
Subject Sample	# 1	0.264	03:24
Breath Volume		1.496	03:24
Air Blank		0.000	03:25
Air Blank		0.000	03:27
Subject Sample	# 2	0.259	03:27
Breath Volume		1.441	03:27
Air Blank		0.000	03:28
Control Test		0.079	03:28
Air Blank		0.000	03:29
Diagnostics Check		OK	03:29

Data Download Date/Time: 03/22/2016 11:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ, MICHAEL, J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: WONGSHUE, ANTHONY
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 02/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 10:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:04

SUBJECT'S LAST NAME: ENGRAM

DATE/TIME OF TEST: 03/19/2016 11:34:31

Diagnostics Check	OK	11:28
Air Blank	0.000	11:28
Control Test	0.080	11:29
Air Blank	0.000	11:29
Subject Sample # 1	0.161	11:30
Breath Volume	2.984	11:30
Air Blank	0.000	11:30
Air Blank	0.000	11:32
Subject Sample # 2	0.161	11:33
Breath Volume	2.316	11:33
Air Blank	0.000	11:33
Control Test	0.081	11:33
Air Blank	0.000	11:34
Diagnostics Check	OK	11:34

Data Download Date/Time: 03/22/2016 11:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,COURTNEY,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MOORE,ANDREW
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 02/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 18:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:09

SUBJECT'S LAST NAME: SERRANO

DATE/TIME OF TEST: 03/19/2016 19:40:45

Diagnostics Check	OK	19:36	
Air Blank	0.000	19:36	
Control Test	0.082	19:37	
Air Blank	0.000	19:37	
Subject Sample	# 1	19:39	Subject Test Refused
Breath Volume	0.000	19:39	
Air Blank	0.000	19:39	
Control Test	0.081	19:40	
Air Blank	0.000	19:40	
Diagnostics Check	OK	19:40	

Data Download Date/Time: 03/22/2016 11:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLEMAN,BRETT,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HAWKINS,DEBORAH
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 02/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:54

SUBJECT'S LAST NAME: ORTIZ MARTINEZ

DATE/TIME OF TEST: 03/20/2016 01:26:28

Diagnostics Check	OK	01:18
Air Blank	0.000	01:19
Control Test	0.081	01:19
Air Blank	0.000	01:20
Subject Sample # 1	0.056	01:21
Breath Volume	1.382	01:21
Air Blank	0.000	01:22
Air Blank	0.000	01:24
Subject Sample # 2	0.060	01:25
Breath Volume	2.160	01:25
Air Blank	0.000	01:25
Control Test	0.082	01:25
Air Blank	0.000	01:26
Diagnostics Check	OK	01:26

Data Download Date/Time: 03/22/2016 11:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ, MICHAEL, J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: WONG SHUE, ANTHONY
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 02/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 11:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:52

SUBJECT'S LAST NAME: NERO

DATE/TIME OF TEST: 03/20/2016 12:23:05

Diagnostics Check	OK	12:17
Air Blank	0.000	12:17
Control Test	0.081	12:18
Air Blank	0.000	12:18
Subject Sample # 1	0.237	12:18
Breath Volume	1.648	12:18
Air Blank	0.000	12:19
Air Blank	0.000	12:21
Subject Sample # 2	0.250	12:21
Breath Volume	1.765	12:21
Air Blank	0.000	12:22
Control Test	0.081	12:22
Air Blank	0.000	12:22
Diagnostics Check	OK	12:23

Data Download Date/Time: 03/22/2016 11:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,COURTNEY,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: GORIS,PATRICIO
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 02/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:01

SUBJECT'S LAST NAME: MOTT

DATE/TIME OF TEST: 03/21/2016 00:32:14

Diagnostics Check	OK	00:25
Air Blank	0.000	00:25
Control Test	0.081	00:26
Air Blank	0.000	00:26
Subject Sample # 1	0.092	00:27
Breath Volume	1.457	00:27
Air Blank	0.000	00:27
Air Blank	0.000	00:29
Subject Sample # 2	0.095	00:30
Breath Volume	1.898	00:30
Air Blank	0.000	00:31
Control Test	0.081	00:31
Air Blank	0.000	00:32
Diagnostics Check	OK	00:32

Data Download Date/Time: 03/22/2016 11:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITFIELD, TIARA, D
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: GALLUP, ELIAS
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 03/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: HICKERSON

DATE/TIME OF TEST: 03/24/2016 01:43:38

Diagnostics Check	OK	01:37
Air Blank	0.000	01:37
Control Test	0.082	01:38
Air Blank	0.000	01:38
Subject Sample # 1	0.131	01:39
Breath Volume	1.605	01:39
Air Blank	0.000	01:39
Air Blank	0.000	01:41
Subject Sample # 2	0.133	01:42
Breath Volume	1.562	01:42
Air Blank	0.000	01:42
Control Test	0.081	01:43
Air Blank	0.000	01:43
Diagnostics Check	OK	01:43

Data Download Date/Time: 03/31/2016 15:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: WEAVER, WILLIAM
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 03/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:38

SUBJECT'S LAST NAME: GANNON

DATE/TIME OF TEST: 03/24/2016 02:04:57

Diagnostics Check		OK	01:58
Air Blank		0.000	01:59
Control Test		0.081	01:59
Air Blank		0.000	02:00
Subject Sample	# 1	0.160	02:00
Breath Volume		2.292	02:00
Air Blank		0.000	02:01
Air Blank		0.000	02:03
Subject Sample	# 2	0.141	02:03
Breath Volume		1.476	02:03
Air Blank		0.000	02:04
Control Test		0.080	02:04
Air Blank		0.000	02:04
Diagnostics Check		OK	02:04

Data Download Date/Time: 03/31/2016 15:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,SHAWN,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DOBBINS,SEAN
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 03/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:12

SUBJECT'S LAST NAME: FAZIODERE

DATE/TIME OF TEST: 03/24/2016 02:47:26

Diagnostics Check		OK	02:35	
Air Blank		0.000	02:36	
Control Test		0.081	02:36	
Air Blank		0.000	02:37	
Subject Sample	# 1	0.157	02:40	Volume Not Met
Breath Volume		0.222	02:40	
Air Blank		0.000	02:40	
Air Blank		0.000	02:42	
Subject Sample	# 2	0.145	02:45	Volume Not Met
Breath Volume		0.351	02:45	
Air Blank		0.000	02:46	
Control Test		0.081	02:46	
Air Blank		0.000	02:47	
Diagnostics Check		OK	02:47	

Data Download Date/Time: 03/31/2016 15:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITFIELD, TIARA, D
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: VAUGHN, JACOB
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 03/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:19

SUBJECT'S LAST NAME: ROBINSON

DATE/TIME OF TEST: 03/24/2016 03:50:27

Diagnostics Check	OK	03:43
Air Blank	0.000	03:44
Control Test	0.082	03:44
Air Blank	0.000	03:44
Subject Sample # 1	0.089	03:45
Breath Volume	3.574	03:45
Air Blank	0.000	03:46
Air Blank	0.000	03:48
Subject Sample # 2	0.088	03:48
Breath Volume	3.890	03:48
Air Blank	0.000	03:49
Control Test	0.081	03:49
Air Blank	0.000	03:50
Diagnostics Check	OK	03:50

Data Download Date/Time: 03/31/2016 15:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITFIELD, TIARA, D
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: GRIMMER, JONATHAN
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 03/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:44

SUBJECT'S LAST NAME: CASTANOLENIS

DATE/TIME OF TEST: 03/25/2016 01:18:08

Diagnostics Check		OK	01:11
Air Blank		0.000	01:12
Control Test		0.081	01:12
Air Blank		0.000	01:12
Subject Sample	# 1	0.200	01:13
Breath Volume		1.367	01:13
Air Blank		0.000	01:13
Air Blank		0.000	01:15
Subject Sample	# 2	0.205	01:16
Breath Volume		1.531	01:16
Air Blank		0.000	01:17
Control Test		0.081	01:17
Air Blank		0.000	01:18
Diagnostics Check		OK	01:18

Data Download Date/Time: 03/31/2016 15:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: PRICE, FORREST
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 03/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:54

SUBJECT'S LAST NAME: FREEMAN

DATE/TIME OF TEST: 03/26/2016 03:30:58

Diagnostics Check		OK	03:17	
Air Blank		0.000	03:18	
Control Test		0.081	03:18	
Air Blank		0.000	03:19	
Subject Sample	# 1	0.127	03:20	
Breath Volume		1.128	03:20	
Air Blank		0.000	03:21	
Air Blank		0.000	03:23	
Subject Sample	# 2	0.129	03:26	Volume Not Met
Breath Volume		0.679	03:26	
Air Blank		0.000	03:27	
Air Blank		0.000	03:28	
Subject Sample	# 3	0.137	03:29	
Breath Volume		1.156	03:29	
Air Blank		0.000	03:30	
Control Test		0.079	03:30	
Air Blank		0.000	03:30	
Diagnostics Check		OK	03:30	

Data Download Date/Time: 03/31/2016 15:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BINET, MARLON
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 03/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:55

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 03/26/2016 22:23:10

Diagnostics Check	OK	22:17
Air Blank	0.000	22:17
Control Test	0.081	22:17
Air Blank	0.000	22:18
Subject Sample # 1	0.105	22:18
Breath Volume	1.535	22:18
Air Blank	0.000	22:19
Air Blank	0.000	22:21
Subject Sample # 2	0.103	22:21
Breath Volume	2.843	22:21
Air Blank	0.000	22:22
Control Test	0.080	22:22
Air Blank	0.000	22:23
Diagnostics Check	OK	22:23

Data Download Date/Time: 03/31/2016 15:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FIGUEROA, HECTOR, A
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: TORRES, NICOLE
ARRESTING OFFICER AGENCY: Windermere PD
LAST AGENCY INSPECTION: 03/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: YING

DATE/TIME OF TEST: 03/27/2016 02:30:10

Diagnostics Check	OK	02:23
Air Blank	0.000	02:23
Control Test	0.080	02:24
Air Blank	0.000	02:24
Subject Sample # 1	0.149	02:25
Breath Volume	2.324	02:25
Air Blank	0.000	02:25
Air Blank	0.000	02:27
Subject Sample # 2	0.153	02:28
Breath Volume	2.300	02:28
Air Blank	0.000	02:29
Control Test	0.081	02:29
Air Blank	0.000	02:30
Diagnostics Check	OK	02:30

Data Download Date/Time: 03/31/2016 15:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FIGUEROA, HECTOR, A
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DOBBINS, SEAN
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 03/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:28

SUBJECT'S LAST NAME: HINTON

DATE/TIME OF TEST: 03/27/2016 03:57:58

Diagnostics Check	OK	03:51
Air Blank	0.000	03:51
Control Test	0.081	03:52
Air Blank	0.000	03:52
Subject Sample # 1	0.147	03:53
Breath Volume	1.621	03:53
Air Blank	0.000	03:53
Air Blank	0.000	03:55
Subject Sample # 2	0.138	03:56
Breath Volume	1.179	03:56
Air Blank	0.000	03:57
Control Test	0.081	03:57
Air Blank	0.000	03:57
Diagnostics Check	OK	03:57

Data Download Date/Time: 03/31/2016 15:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCREWS,DARCY,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ZITO,ANTHONY
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 03/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:48

SUBJECT'S LAST NAME: GALLO

DATE/TIME OF TEST: 03/27/2016 04:17:54

Diagnostics Check	OK	04:11
Air Blank	0.000	04:11
Control Test	0.082	04:12
Air Blank	0.000	04:12
Subject Sample # 1	0.110	04:13
Breath Volume	2.402	04:13
Air Blank	0.000	04:14
Air Blank	0.000	04:16
Subject Sample # 2	0.111	04:16
Breath Volume	2.785	04:16
Air Blank	0.000	04:17
Control Test	0.081	04:17
Air Blank	0.000	04:17
Diagnostics Check	OK	04:17

Data Download Date/Time: 03/31/2016 15:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FIGUEROA, HECTOR, A
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BINET, MARLON
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 03/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:01

SUBJECT'S LAST NAME: HECKERMAN

DATE/TIME OF TEST: 03/27/2016 05:32:40

Diagnostics Check		OK	05:26
Air Blank		0.000	05:26
Control Test		0.080	05:27
Air Blank		0.000	05:27
Subject Sample	# 1	0.182	05:28
Breath Volume		2.351	05:28
Air Blank		0.000	05:28
Air Blank		0.000	05:30
Subject Sample	# 2	0.178	05:31
Breath Volume		1.347	05:31
Air Blank		0.000	05:31
Control Test		0.080	05:32
Air Blank		0.000	05:32
Diagnostics Check		OK	05:32

Data Download Date/Time: 03/31/2016 15:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITFIELD, TIARA, D
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: GAUNTLETT, BEN
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 03/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: SAUDARGAS

DATE/TIME OF TEST: 03/27/2016 23:29:02

Diagnostics Check	OK	23:25	
Air Blank	0.000	23:26	
Control Test	0.082	23:26	
Air Blank	0.000	23:27	
Subject Sample	# 1	23:27	Subject Test Refused
Breath Volume	0.000	23:27	
Air Blank	0.000	23:28	
Control Test	0.081	23:28	
Air Blank	0.000	23:28	
Diagnostics Check	OK	23:29	

Data Download Date/Time: 03/31/2016 15:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: PRINCE, BLAKE
ARRESTING OFFICER AGENCY: Ocoee PD
LAST AGENCY INSPECTION: 03/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:49

SUBJECT'S LAST NAME: PEREZ AUGUSTIN

DATE/TIME OF TEST: 03/28/2016 00:18:39

Diagnostics Check	OK	00:12
Air Blank	0.000	00:12
Control Test	0.082	00:12
Air Blank	0.000	00:13
Subject Sample # 1	0.166	00:14
Breath Volume	1.792	00:14
Air Blank	0.000	00:14
Air Blank	0.000	00:16
Subject Sample # 2	0.174	00:17
Breath Volume	2.300	00:17
Air Blank	0.000	00:17
Control Test	0.080	00:18
Air Blank	0.000	00:18
Diagnostics Check	OK	00:18

Data Download Date/Time: 03/31/2016 15:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITFIELD, TIARA, D
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ALEJANDRO, GENESIS
ARRESTING OFFICER AGENCY: Winter Park PD
LAST AGENCY INSPECTION: 03/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: ISWED

DATE/TIME OF TEST: 03/28/2016 01:02:47

Diagnostics Check	OK	00:59	
Air Blank	0.000	00:59	
Control Test	0.081	00:59	
Air Blank	0.000	01:00	
Subject Sample	# 1	01:01	Subject Test Refused
Breath Volume	0.000	01:01	
Air Blank	0.000	01:01	
Control Test	0.080	01:02	
Air Blank	0.000	01:02	
Diagnostics Check	OK	01:02	

Data Download Date/Time: 03/31/2016 15:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: EL KHOURY, ELIAS
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 03/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:02

SUBJECT'S LAST NAME: ROSENBERG

DATE/TIME OF TEST: 03/28/2016 01:34:31

Diagnostics Check	OK	01:23	
Air Blank	0.000	01:23	
Control Test	0.081	01:23	
Air Blank	0.000	01:24	
Subject Sample	# 1	01:27	No Sample Provided
Breath Volume	0.000	01:27	
Air Blank	0.000	01:28	
Air Blank	0.000	01:30	
Subject Sample	# 2	01:33	No Sample Provided
Breath Volume	0.000	01:33	
Air Blank	0.000	01:33	
Control Test	0.080	01:33	
Air Blank	0.000	01:34	
Diagnostics Check	OK	01:34	

Data Download Date/Time: 03/31/2016 15:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITFIELD, TIARA, D
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: TORRES, NICOLE
ARRESTING OFFICER AGENCY: Windermere PD
LAST AGENCY INSPECTION: 03/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:43

SUBJECT'S LAST NAME: GORDON

DATE/TIME OF TEST: 03/28/2016 04:16:06

Diagnostics Check	OK	04:09
Air Blank	0.000	04:09
Control Test	0.080	04:10
Air Blank	0.000	04:10
Subject Sample # 1	0.155	04:11
Breath Volume	1.273	04:11
Air Blank	0.000	04:11
Air Blank	0.000	04:13
Subject Sample # 2	0.157	04:14
Breath Volume	3.292	04:14
Air Blank	0.000	04:15
Control Test	0.080	04:15
Air Blank	0.000	04:16
Diagnostics Check	OK	04:16

Data Download Date/Time: 03/31/2016 15:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: URIBE, SERGIO
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 03/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:08

SUBJECT'S LAST NAME: KERBY

DATE/TIME OF TEST: 03/28/2016 23:35:21

Diagnostics Check	OK	23:29
Air Blank	0.000	23:29
Control Test	0.081	23:30
Air Blank	0.000	23:30
Subject Sample # 1	0.175	23:30
Breath Volume	1.550	23:30
Air Blank	0.000	23:31
Air Blank	0.000	23:33
Subject Sample # 2	0.185	23:33
Breath Volume	1.507	23:33
Air Blank	0.000	23:34
Control Test	0.081	23:34
Air Blank	0.000	23:35
Diagnostics Check	OK	23:35

Data Download Date/Time: 03/31/2016 15:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITFIELD, TIARA, D
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: KEANE, PETER
ARRESTING OFFICER AGENCY: Winter Garden PD
LAST AGENCY INSPECTION: 03/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:58

SUBJECT'S LAST NAME: MCCARTHY

DATE/TIME OF TEST: 03/29/2016 03:28:12

Diagnostics Check		OK	03:22
Air Blank		0.000	03:22
Control Test		0.081	03:23
Air Blank		0.000	03:23
Subject Sample	# 1	0.248	03:23
Breath Volume		1.507	03:23
Air Blank		0.000	03:24
Air Blank		0.000	03:26
Subject Sample	# 2	0.249	03:26
Breath Volume		1.535	03:26
Air Blank		0.000	03:27
Control Test		0.081	03:27
Air Blank		0.000	03:28
Diagnostics Check		OK	03:28

Data Download Date/Time: 03/31/2016 15:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,COURTNEY,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: URIBE,SERGIO
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 03/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:18

SUBJECT'S LAST NAME: PINOCHET

DATE/TIME OF TEST: 03/29/2016 03:47:24

Diagnostics Check	OK	03:41
Air Blank	0.000	03:42
Control Test	0.082	03:42
Air Blank	0.000	03:42
Subject Sample # 1	0.028	03:43
Breath Volume	1.917	03:43
Air Blank	0.000	03:43
Air Blank	0.000	03:45
Subject Sample # 2	0.028	03:45
Breath Volume	1.523	03:45
Air Blank	0.000	03:46
Control Test	0.082	03:46
Air Blank	0.000	03:47
Diagnostics Check	OK	03:47

Data Download Date/Time: 03/31/2016 15:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FIGUEROA, HECTOR, A
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: KRALICH, BRENDEN
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 03/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: WILKINS

DATE/TIME OF TEST: 03/31/2016 01:53:52

Diagnostics Check	OK	01:47
Air Blank	0.000	01:47
Control Test	0.081	01:47
Air Blank	0.000	01:48
Subject Sample # 1	0.174	01:49
Breath Volume	2.722	01:49
Air Blank	0.000	01:49
Air Blank	0.000	01:51
Subject Sample # 2	0.165	01:52
Breath Volume	2.398	01:52
Air Blank	0.000	01:52
Control Test	0.081	01:53
Air Blank	0.000	01:53
Diagnostics Check	OK	01:53

Data Download Date/Time: 03/31/2016 15:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITFIELD, TIARA, D
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: VAUGHN, JACOB
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 03/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:48

SUBJECT'S LAST NAME: CONCEPCION

DATE/TIME OF TEST: 03/31/2016 03:18:52

Diagnostics Check	OK	03:12
Air Blank	0.000	03:13
Control Test	0.082	03:13
Air Blank	0.000	03:14
Subject Sample # 1	0.127	03:14
Breath Volume	1.488	03:14
Air Blank	0.000	03:15
Air Blank	0.000	03:17
Subject Sample # 2	0.127	03:17
Breath Volume	1.125	03:17
Air Blank	0.000	03:17
Control Test	0.081	03:18
Air Blank	0.000	03:18
Diagnostics Check	OK	03:18

Data Download Date/Time: 03/31/2016 15:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,SHAWN,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CHARLEBOIS,MYLES
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 03/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 05:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:47

SUBJECT'S LAST NAME: JEAN BAPTISTE

DATE/TIME OF TEST: 03/31/2016 06:23:32

Diagnostics Check		OK	06:11	
Air Blank		0.000	06:12	
Control Test		0.082	06:12	
Air Blank		0.000	06:13	
Subject Sample	# 1	0.091	06:16	Volume Not Met
Breath Volume		0.347	06:16	
Air Blank		0.000	06:16	
Air Blank		0.000	06:18	
Subject Sample	# 2	0.088	06:22	Volume Not Met
Breath Volume		0.890	06:22	
Air Blank		0.000	06:22	
Control Test		0.081	06:22	
Air Blank		0.000	06:23	
Diagnostics Check		OK	06:23	

Data Download Date/Time: 03/31/2016 15:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,SHAWN,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CHARLEBOIS,MYLES
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 03/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 05:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:47

SUBJECT'S LAST NAME: JEAN BAPTISTE

DATE/TIME OF TEST: 03/31/2016 06:31:19

Diagnostics Check	OK	06:24
Air Blank	0.000	06:25
Control Test	0.081	06:25
Air Blank	0.000	06:25
Subject Sample # 1	0.107	06:26
Breath Volume	2.132	06:26
Air Blank	0.000	06:26
Air Blank	0.000	06:28
Subject Sample # 2	0.111	06:29
Breath Volume	1.769	06:29
Air Blank	0.000	06:30
Control Test	0.081	06:30
Air Blank	0.000	06:31
Diagnostics Check	OK	06:31

Data Download Date/Time: 03/31/2016 15:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FIGUEROA, HECTOR, A
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: KEANE, PETER
ARRESTING OFFICER AGENCY: Winter Garden PD
LAST AGENCY INSPECTION: 02/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:43

SUBJECT'S LAST NAME: REYNOLDS

DATE/TIME OF TEST: 03/01/2016 01:19:51

Diagnostics Check		OK	01:05	
Air Blank		0.000	01:05	
Control Test		0.083	01:05	
Air Blank		0.000	01:06	
Subject Sample	# 1	0.115	01:09	Volume Not Met
Breath Volume		0.250	01:09	
Air Blank		0.000	01:10	
Air Blank		0.000	01:11	
Subject Sample	# 2	0.145	01:12	
Breath Volume		1.460	01:12	
Air Blank		0.000	01:13	
Air Blank		0.000	01:15	
Subject Sample	# 3	0.134	01:18	
Breath Volume		1.121	01:18	
Air Blank		0.000	01:18	
Control Test		0.081	01:19	
Air Blank		0.000	01:19	
Diagnostics Check		OK	01:19	

Data Download Date/Time: 03/03/2016 15:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: WEAVER, WILLIAM
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 02/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: DEL VALLE VARGAS

DATE/TIME OF TEST: 03/01/2016 03:12:40

Diagnostics Check	OK	03:08	
Air Blank	0.000	03:09	
Control Test	0.082	03:09	
Air Blank	0.000	03:10	
Subject Sample	# 1	03:11	Subject Test Refused
Breath Volume	0.000	03:11	
Air Blank	0.000	03:11	
Control Test	0.082	03:12	
Air Blank	0.000	03:12	
Diagnostics Check	OK	03:12	

Data Download Date/Time: 03/03/2016 15:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PADRON,MARCIE,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: JONES,EUGENE
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 02/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 11:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:20

SUBJECT'S LAST NAME: BRAGG

DATE/TIME OF TEST: 03/01/2016 13:01:40

Diagnostics Check		OK	12:47	
Air Blank		0.000	12:47	
Control Test		0.083	12:47	
Air Blank		0.000	12:48	
Subject Sample	# 1	0.000	12:51	Volume Not Met
Breath Volume		0.457	12:51	
Air Blank		0.000	12:52	
Air Blank		0.000	12:54	
Subject Sample	# 2	0.000	12:54	
Breath Volume		1.242	12:54	
Air Blank		0.000	12:55	
Air Blank		0.000	12:57	
Subject Sample	# 3	0.000	13:00	Volume Not Met
Breath Volume		0.625	13:00	
Air Blank		0.000	13:00	
Control Test		0.083	13:01	
Air Blank		0.000	13:01	
Diagnostics Check		OK	13:01	

Data Download Date/Time: 03/03/2016 15:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITFIELD, TIARA, D
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: EVANS, JOSHUA
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 02/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:52

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 03/03/2016 01:22:17

Diagnostics Check		OK	01:15
Air Blank		0.000	01:16
Control Test		0.082	01:16
Air Blank		0.000	01:17
Subject Sample	# 1	0.136	01:17
Breath Volume		1.687	01:17
Air Blank		0.000	01:18
Air Blank		0.000	01:20
Subject Sample	# 2	0.126	01:20
Breath Volume		1.414	01:20
Air Blank		0.000	01:21
Control Test		0.083	01:21
Air Blank		0.000	01:22
Diagnostics Check		OK	01:22

Data Download Date/Time: 03/03/2016 15:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,SHAWN,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: THORNTON,BRANDEN
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 02/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:08

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 03/03/2016 02:39:34

Diagnostics Check	OK	02:31
Air Blank	0.000	02:31
Control Test	0.082	02:31
Air Blank	0.000	02:32
Subject Sample # 1	0.131	02:34
Breath Volume	1.796	02:34
Air Blank	0.000	02:35
Air Blank	0.000	02:37
Subject Sample # 2	0.132	02:37
Breath Volume	2.750	02:37
Air Blank	0.000	02:38
Control Test	0.081	02:38
Air Blank	0.000	02:39
Diagnostics Check	OK	02:39

Data Download Date/Time: 03/03/2016 15:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITFIELD, TIARA, D
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DOBBINS, SEAN
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 02/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: MEDELUS

DATE/TIME OF TEST: 03/03/2016 03:42:32

Diagnostics Check	OK	03:34
Air Blank	0.000	03:35
Control Test	0.082	03:35
Air Blank	0.000	03:36
Subject Sample # 1	0.198	03:38
Breath Volume	1.718	03:38
Air Blank	0.000	03:38
Air Blank	0.000	03:40
Subject Sample # 2	0.203	03:40
Breath Volume	1.375	03:40
Air Blank	0.000	03:41
Control Test	0.081	03:41
Air Blank	0.000	03:42
Diagnostics Check	OK	03:42

Data Download Date/Time: 03/03/2016 15:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITFIELD, TIARA, D
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MOORE, CHRISTOPHER
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 02/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: BIERKAN

DATE/TIME OF TEST: 03/04/2016 01:14:55

Diagnostics Check		OK	01:08
Air Blank		0.000	01:09
Control Test		0.081	01:09
Air Blank		0.000	01:10
Subject Sample	# 1	0.200	01:10
Breath Volume		1.558	01:10
Air Blank		0.000	01:11
Air Blank		0.000	01:13
Subject Sample	# 2	0.198	01:13
Breath Volume		1.488	01:13
Air Blank		0.000	01:14
Control Test		0.082	01:14
Air Blank		0.000	01:14
Diagnostics Check		OK	01:14

Data Download Date/Time: 03/10/2016 16:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON,TRAVIS,W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DANJOU,SCOTT
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 02/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:07

SUBJECT'S LAST NAME: COHEE

DATE/TIME OF TEST: 03/04/2016 01:33:42

Diagnostics Check	OK	01:27
Air Blank	0.000	01:27
Control Test	0.082	01:28
Air Blank	0.000	01:28
Subject Sample # 1	0.194	01:29
Breath Volume	1.527	01:29
Air Blank	0.000	01:29
Air Blank	0.000	01:31
Subject Sample # 2	0.194	01:32
Breath Volume	1.156	01:32
Air Blank	0.000	01:32
Control Test	0.081	01:33
Air Blank	0.000	01:33
Diagnostics Check	OK	01:33

Data Download Date/Time: 03/10/2016 16:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,SHAWN,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DOBBINS,SEAN
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 02/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:27

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 03/04/2016 03:57:22

Diagnostics Check	OK	03:50
Air Blank	0.000	03:51
Control Test	0.082	03:51
Air Blank	0.000	03:52
Subject Sample # 1	0.096	03:52
Breath Volume	1.867	03:52
Air Blank	0.000	03:53
Air Blank	0.000	03:55
Subject Sample # 2	0.091	03:55
Breath Volume	1.464	03:55
Air Blank	0.000	03:56
Control Test	0.082	03:56
Air Blank	0.000	03:57
Diagnostics Check	OK	03:57

Data Download Date/Time: 03/10/2016 16:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HALLOCK, STEPHEN
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 02/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:56

SUBJECT'S LAST NAME: CAMUY VEGA

DATE/TIME OF TEST: 03/04/2016 04:23:22

Diagnostics Check	OK	04:16
Air Blank	0.000	04:17
Control Test	0.083	04:17
Air Blank	0.000	04:17
Subject Sample # 1	0.125	04:18
Breath Volume	2.753	04:18
Air Blank	0.000	04:19
Air Blank	0.000	04:21
Subject Sample # 2	0.122	04:21
Breath Volume	3.109	04:21
Air Blank	0.000	04:22
Control Test	0.081	04:22
Air Blank	0.000	04:23
Diagnostics Check	OK	04:23

Data Download Date/Time: 03/10/2016 16:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITFIELD, TIARA, D
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HAWKINS, DEBORAH
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 02/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 03/04/2016 21:56:43

Diagnostics Check		OK	21:44	
Air Blank		0.000	21:44	
Control Test		0.083	21:45	
Air Blank		0.000	21:45	
Subject Sample	# 1	0.148	21:48	Volume Not Met
Breath Volume		0.714	21:48	
Air Blank		0.000	21:49	
Air Blank		0.000	21:51	
Subject Sample	# 2	0.153	21:52	
Breath Volume		2.250	21:52	
Air Blank		0.000	21:52	
Air Blank		0.000	21:54	
Subject Sample	# 3	0.154	21:55	
Breath Volume		1.507	21:55	
Air Blank		0.000	21:55	
Control Test		0.080	21:56	
Air Blank		0.000	21:56	
Diagnostics Check		OK	21:56	

Data Download Date/Time: 03/10/2016 16:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITFIELD, TIARA, D
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: KEANE, PETER
ARRESTING OFFICER AGENCY: Winter Garden PD
LAST AGENCY INSPECTION: 02/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: CONROY

DATE/TIME OF TEST: 03/05/2016 00:53:24

Diagnostics Check	OK	00:44
Air Blank	0.000	00:45
Control Test	0.082	00:45
Air Blank	0.000	00:45
Subject Sample # 1	0.220	00:48
Breath Volume	1.218	00:48
Air Blank	0.000	00:49
Air Blank	0.000	00:50
Subject Sample # 2	0.220	00:51
Breath Volume	1.449	00:51
Air Blank	0.000	00:52
Control Test	0.080	00:52
Air Blank	0.000	00:53
Diagnostics Check	OK	00:53

Data Download Date/Time: 03/10/2016 16:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITFIELD, TIARA, D
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MACDONALD, JOHN
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 02/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:32

SUBJECT'S LAST NAME: QUARTERMAN

DATE/TIME OF TEST: 03/05/2016 03:02:51

Diagnostics Check	OK	02:56
Air Blank	0.000	02:57
Control Test	0.082	02:57
Air Blank	0.000	02:57
Subject Sample # 1	0.110	02:58
Breath Volume	3.894	02:58
Air Blank	0.000	02:59
Air Blank	0.000	03:00
Subject Sample # 2	0.108	03:01
Breath Volume	3.281	03:01
Air Blank	0.000	03:01
Control Test	0.082	03:02
Air Blank	0.000	03:02
Diagnostics Check	OK	03:02

Data Download Date/Time: 03/10/2016 16:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITFIELD, TIARA, D
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HALLOCK, STEPHEN
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 02/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: SUREZRAMIREZ

DATE/TIME OF TEST: 03/05/2016 04:02:55

Diagnostics Check	OK	03:54
Air Blank	0.000	03:54
Control Test	0.082	03:55
Air Blank	0.000	03:55
Subject Sample # 1	0.153	03:57
Breath Volume	1.804	03:57
Air Blank	0.000	03:58
Air Blank	0.000	03:59
Subject Sample # 2	0.137	04:01
Breath Volume	1.375	04:01
Air Blank	0.000	04:02
Control Test	0.080	04:02
Air Blank	0.000	04:02
Diagnostics Check	OK	04:02

Data Download Date/Time: 03/10/2016 16:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITFIELD, TIARA, D
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HERNANDEZ, FRANK
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 02/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:06

SUBJECT'S LAST NAME: ROMERO

DATE/TIME OF TEST: 03/05/2016 05:43:36

Diagnostics Check		OK	05:31	
Air Blank		0.000	05:32	
Control Test		0.081	05:32	
Air Blank		0.000	05:33	
Subject Sample	# 1	0.169	05:36	Volume Not Met
Breath Volume		0.570	05:36	
Air Blank		0.000	05:37	
Air Blank		0.000	05:38	
Subject Sample	# 2	0.211	05:41	Volume Not Met
Breath Volume		0.253	05:41	
Air Blank		0.000	05:42	
Control Test		0.081	05:42	
Air Blank		0.000	05:43	
Diagnostics Check		OK	05:43	

Data Download Date/Time: 03/10/2016 16:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITFIELD, TIARA, D
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: WINBURN, TAYLOR
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 02/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:19

SUBJECT'S LAST NAME: MORENO

DATE/TIME OF TEST: 03/06/2016 01:53:24

Diagnostics Check	OK	01:43
Air Blank	0.000	01:43
Control Test	0.083	01:43
Air Blank	0.000	01:44
Subject Sample # 1	0.227	01:44
Breath Volume	1.988	01:44
Air Blank	0.000	01:45
Air Blank	0.000	01:47
Subject Sample # 2	0.259	01:47
Breath Volume	1.863	01:47
Air Blank	0.000	01:48
Air Blank	0.000	01:50
Subject Sample # 3	0.220	01:51
Breath Volume	1.785	01:51
Air Blank	0.000	01:52
Control Test	0.082	01:52
Air Blank	0.000	01:53
Diagnostics Check	OK	01:53

Data Download Date/Time: 03/10/2016 16:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITFIELD, TIARA, D
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: VAUGHN, JACOB
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 02/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: SAPP

DATE/TIME OF TEST: 03/06/2016 03:35:36

Diagnostics Check	OK	03:28
Air Blank	0.000	03:28
Control Test	0.081	03:29
Air Blank	0.000	03:29
Subject Sample # 1	0.081	03:30
Breath Volume	2.222	03:30
Air Blank	0.000	03:31
Air Blank	0.000	03:33
Subject Sample # 2	0.083	03:34
Breath Volume	2.730	03:34
Air Blank	0.000	03:34
Control Test	0.082	03:34
Air Blank	0.000	03:35
Diagnostics Check	OK	03:35

Data Download Date/Time: 03/10/2016 16:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON,TRAVIS,W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MATHISEN,LEIGH
ARRESTING OFFICER AGENCY: Winter Garden PD
LAST AGENCY INSPECTION: 02/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:37

SUBJECT'S LAST NAME: ALVAREZ

DATE/TIME OF TEST: 03/06/2016 04:04:30

Diagnostics Check	OK	03:58
Air Blank	0.000	03:59
Control Test	0.082	03:59
Air Blank	0.000	03:59
Subject Sample # 1	0.000	04:00
Breath Volume	2.183	04:00
Air Blank	0.000	04:00
Air Blank	0.000	04:02
Subject Sample # 2	0.000	04:03
Breath Volume	2.027	04:03
Air Blank	0.000	04:03
Control Test	0.081	04:03
Air Blank	0.000	04:04
Diagnostics Check	OK	04:04

Data Download Date/Time: 03/10/2016 16:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FIGUEROA, HECTOR, A
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HUTTO, GERALD
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 02/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:45

SUBJECT'S LAST NAME: RIVAS

DATE/TIME OF TEST: 03/06/2016 05:14:59

Diagnostics Check	OK	05:08
Air Blank	0.000	05:08
Control Test	0.082	05:08
Air Blank	0.000	05:09
Subject Sample # 1	0.157	05:09
Breath Volume	1.496	05:09
Air Blank	0.000	05:10
Air Blank	0.000	05:12
Subject Sample # 2	0.159	05:13
Breath Volume	3.011	05:13
Air Blank	0.000	05:14
Control Test	0.082	05:14
Air Blank	0.000	05:14
Diagnostics Check	OK	05:14

Data Download Date/Time: 03/10/2016 16:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: WONG SHUR, ANTHONY
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 02/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 06:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:38

SUBJECT'S LAST NAME: HENRY

DATE/TIME OF TEST: 03/06/2016 07:09:51

Diagnostics Check	OK	07:02
Air Blank	0.000	07:03
Control Test	0.082	07:03
Air Blank	0.000	07:04
Subject Sample # 1	0.167	07:05
Breath Volume	1.207	07:05
Air Blank	0.000	07:05
Air Blank	0.000	07:07
Subject Sample # 2	0.165	07:08
Breath Volume	1.691	07:08
Air Blank	0.000	07:08
Control Test	0.082	07:09
Air Blank	0.000	07:09
Diagnostics Check	OK	07:09

Data Download Date/Time: 03/10/2016 16:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FIGUEROA,HECTOR,A
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: KNIGHT,RYAN
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 02/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:09

SUBJECT'S LAST NAME: NIVAR

DATE/TIME OF TEST: 03/07/2016 02:43:50

Diagnostics Check	OK	02:36
Air Blank	0.000	02:36
Control Test	0.082	02:36
Air Blank	0.000	02:37
Subject Sample # 1	0.267	02:38
Breath Volume	2.722	02:38
Air Blank	0.000	02:38
Air Blank	0.000	02:40
Subject Sample # 2	0.250	02:42
Breath Volume	1.933	02:42
Air Blank	0.000	02:42
Control Test	0.081	02:43
Air Blank	0.000	02:43
Diagnostics Check	OK	02:43

Data Download Date/Time: 03/10/2016 16:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FIGUEROA, HECTOR, A
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: URIBE, SERGIO
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 02/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:38

SUBJECT'S LAST NAME: AVILES VIERA

DATE/TIME OF TEST: 03/07/2016 04:07:54

Diagnostics Check	OK	04:00
Air Blank	0.000	04:01
Control Test	0.082	04:01
Air Blank	0.000	04:02
Subject Sample # 1	0.089	04:03
Breath Volume	1.445	04:03
Air Blank	0.000	04:04
Air Blank	0.000	04:06
Subject Sample # 2	0.083	04:06
Breath Volume	1.476	04:06
Air Blank	0.000	04:06
Control Test	0.081	04:07
Air Blank	0.000	04:07
Diagnostics Check	OK	04:07

Data Download Date/Time: 03/10/2016 16:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FIGUEROA, HECTOR, A
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: WALKER, JUSTIN
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 02/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:21

SUBJECT'S LAST NAME: REVOLORIO

DATE/TIME OF TEST: 03/09/2016 03:46:30

Diagnostics Check	OK	03:43	
Air Blank	0.000	03:44	
Control Test	0.083	03:44	
Air Blank	0.000	03:44	
Subject Sample	# 1	03:45	Subject Test Refused
Breath Volume	0.000	03:45	
Air Blank	0.000	03:45	
Control Test	0.082	03:45	
Air Blank	0.000	03:46	
Diagnostics Check	OK	03:46	

Data Download Date/Time: 03/10/2016 16:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITFIELD, TIARA, D
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HAIGHT, MONTAGUE
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 02/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:23

SUBJECT'S LAST NAME: FLOOD

DATE/TIME OF TEST: 03/09/2016 23:52:46

Diagnostics Check	OK	23:46
Air Blank	0.000	23:47
Control Test	0.082	23:47
Air Blank	0.000	23:48
Subject Sample # 1	0.354	23:48
Breath Volume	2.406	23:48
Air Blank	0.000	23:49
Air Blank	0.000	23:50
Subject Sample # 2	0.352	23:51
Breath Volume	2.410	23:51
Air Blank	0.000	23:51
Control Test	0.081	23:52
Air Blank	0.000	23:52
Diagnostics Check	OK	23:52

Data Download Date/Time: 03/10/2016 16:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HUTTO, GERLAD
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 02/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:27

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 03/10/2016 02:01:48

Diagnostics Check		OK	01:49	
Air Blank		0.000	01:49	
Control Test		0.081	01:49	
Air Blank		0.000	01:50	
Subject Sample	# 1		01:53	No Sample Provided
Breath Volume		0.000	01:53	
Air Blank		0.000	01:53	
Air Blank		0.000	01:55	
Subject Sample	# 2	0.231	01:57	
Breath Volume		2.695	01:57	
Air Blank		0.000	01:57	
Air Blank		0.000	01:59	
Subject Sample	# 3	0.213	02:00	
Breath Volume		1.320	02:00	
Air Blank		0.000	02:00	
Control Test		0.082	02:01	
Air Blank		0.000	02:01	
Diagnostics Check		OK	02:01	

Data Download Date/Time: 03/10/2016 16:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITFIELD, TIARA, D
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CARDINAL, TIMOTHY
ARRESTING OFFICER AGENCY: Edgewood PD
LAST AGENCY INSPECTION: 02/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: TECCHIOPASSOS

DATE/TIME OF TEST: 03/10/2016 02:42:03

Diagnostics Check		OK	02:35
Air Blank		0.000	02:36
Control Test		0.082	02:36
Air Blank		0.000	02:37
Subject Sample	# 1	0.201	02:37
Breath Volume		2.003	02:37
Air Blank		0.000	02:38
Air Blank		0.000	02:40
Subject Sample	# 2	0.188	02:40
Breath Volume		1.562	02:40
Air Blank		0.000	02:41
Control Test		0.081	02:41
Air Blank		0.000	02:41
Diagnostics Check		OK	02:42

Data Download Date/Time: 03/10/2016 16:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITFIELD, TIARA, D
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HAIGHT, MONTAGUE
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 02/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:01

SUBJECT'S LAST NAME: MORALES GUERRERO

DATE/TIME OF TEST: 03/11/2016 00:33:23

Diagnostics Check	OK	00:27
Air Blank	0.000	00:27
Control Test	0.082	00:27
Air Blank	0.000	00:28
Subject Sample # 1	0.177	00:28
Breath Volume	2.609	00:28
Air Blank	0.000	00:29
Air Blank	0.000	00:31
Subject Sample # 2	0.177	00:31
Breath Volume	3.160	00:31
Air Blank	0.000	00:32
Control Test	0.081	00:32
Air Blank	0.000	00:33
Diagnostics Check	OK	00:33

Data Download Date/Time: 03/17/2016 16:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: GRIMMER, JONATHAN
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 02/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:29

SUBJECT'S LAST NAME: NUNEZ SANTOS

DATE/TIME OF TEST: 03/11/2016 00:58:42

Diagnostics Check	OK	00:51
Air Blank	0.000	00:52
Control Test	0.082	00:52
Air Blank	0.000	00:53
Subject Sample # 1	0.000	00:53
Breath Volume	1.191	00:53
Air Blank	0.000	00:53
Air Blank	0.000	00:55
Subject Sample # 2	0.000	00:57
Breath Volume	1.597	00:57
Air Blank	0.000	00:57
Control Test	0.082	00:58
Air Blank	0.000	00:58
Diagnostics Check	OK	00:58

Data Download Date/Time: 03/17/2016 16:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITFIELD, TIARA, D
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: FELLOWS, BRENT
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 02/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:18

SUBJECT'S LAST NAME: AZROUR

DATE/TIME OF TEST: 03/11/2016 01:55:17

Diagnostics Check	OK	01:43	
Air Blank	0.000	01:44	
Control Test	0.081	01:44	
Air Blank	0.000	01:45	
Subject Sample	# 1	01:48	No Sample Provided
Breath Volume	0.000	01:48	
Air Blank	0.000	01:48	
Air Blank	0.000	01:50	
Subject Sample	# 2	01:53	No Sample Provided
Breath Volume	0.000	01:53	
Air Blank	0.000	01:54	
Control Test	0.082	01:54	
Air Blank	0.000	01:55	
Diagnostics Check	OK	01:55	

Data Download Date/Time: 03/17/2016 16:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITFIELD, TIARA, D
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: KIPP, ALEXANDER
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 02/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:48

SUBJECT'S LAST NAME: SPRINGER

DATE/TIME OF TEST: 03/11/2016 03:20:43

Diagnostics Check	OK	03:13
Air Blank	0.000	03:13
Control Test	0.082	03:14
Air Blank	0.000	03:14
Subject Sample # 1	0.228	03:16
Breath Volume	2.148	03:16
Air Blank	0.000	03:16
Air Blank	0.000	03:18
Subject Sample # 2	0.226	03:19
Breath Volume	2.046	03:19
Air Blank	0.000	03:19
Control Test	0.081	03:20
Air Blank	0.000	03:20
Diagnostics Check	OK	03:20

Data Download Date/Time: 03/17/2016 16:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: URIBE, SERGIO
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 02/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:09

SUBJECT'S LAST NAME: ANTHONY

DATE/TIME OF TEST: 03/11/2016 03:42:58

Diagnostics Check	OK	03:36
Air Blank	0.000	03:37
Control Test	0.082	03:37
Air Blank	0.000	03:37
Subject Sample # 1	0.236	03:38
Breath Volume	3.179	03:38
Air Blank	0.000	03:39
Air Blank	0.000	03:40
Subject Sample # 2	0.237	03:41
Breath Volume	2.929	03:41
Air Blank	0.000	03:42
Control Test	0.081	03:42
Air Blank	0.000	03:42
Diagnostics Check	OK	03:42

Data Download Date/Time: 03/17/2016 16:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITFIELD, TIARA, D
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: RIVERA, LUIS
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 02/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: STEELE

DATE/TIME OF TEST: 03/11/2016 04:10:22

Diagnostics Check		OK	04:04
Air Blank		0.000	04:04
Control Test		0.082	04:04
Air Blank		0.000	04:05
Subject Sample	# 1	0.287	04:05
Breath Volume		1.921	04:05
Air Blank		0.000	04:06
Air Blank		0.000	04:08
Subject Sample	# 2	0.279	04:08
Breath Volume		1.921	04:08
Air Blank		0.000	04:09
Control Test		0.081	04:09
Air Blank		0.000	04:10
Diagnostics Check		OK	04:10

Data Download Date/Time: 03/17/2016 16:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITFIELD, TIARA, D
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CAMPBELL, ROBERT
ARRESTING OFFICER AGENCY: Apopka PD
LAST AGENCY INSPECTION: 02/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:48

SUBJECT'S LAST NAME: LEA

DATE/TIME OF TEST: 03/11/2016 23:18:27

Diagnostics Check	OK	23:12
Air Blank	0.000	23:12
Control Test	0.082	23:12
Air Blank	0.000	23:13
Subject Sample # 1	0.000	23:14
Breath Volume	3.347	23:14
Air Blank	0.000	23:14
Air Blank	0.000	23:16
Subject Sample # 2	0.000	23:17
Breath Volume	3.351	23:17
Air Blank	0.000	23:17
Control Test	0.082	23:17
Air Blank	0.000	23:18
Diagnostics Check	OK	23:18

Data Download Date/Time: 03/17/2016 16:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,SHAWN,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ELLER,ASHLEY
ARRESTING OFFICER AGENCY: Apopka PD
LAST AGENCY INSPECTION: 02/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:54

SUBJECT'S LAST NAME: MORALES

DATE/TIME OF TEST: 03/12/2016 00:24:30

Diagnostics Check	OK	00:18
Air Blank	0.000	00:18
Control Test	0.082	00:19
Air Blank	0.000	00:19
Subject Sample # 1	0.017	00:20
Breath Volume	2.933	00:20
Air Blank	0.000	00:20
Air Blank	0.000	00:22
Subject Sample # 2	0.015	00:23
Breath Volume	3.234	00:23
Air Blank	0.000	00:23
Control Test	0.083	00:23
Air Blank	0.000	00:24
Diagnostics Check	OK	00:24

Data Download Date/Time: 03/17/2016 16:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITFIELD, TIARA, D
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: PRINCE, BLAKE
ARRESTING OFFICER AGENCY: Ocoee PD
LAST AGENCY INSPECTION: 02/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:42

SUBJECT'S LAST NAME: KECLIK

DATE/TIME OF TEST: 03/12/2016 01:17:50

Diagnostics Check	OK	01:06		
Air Blank	0.000	01:06		
Control Test	0.081	01:07		
Air Blank	0.000	01:07		
Subject Sample	# 1	01:10	No Sample Provided	
Breath Volume	0.000	01:10		
Air Blank	0.000	01:11		
Air Blank	0.000	01:13		
Subject Sample	# 2	0.103	01:16	Volume Not Met
Breath Volume	0.019	01:16		
Air Blank	0.000	01:16		
Control Test	0.082	01:17		
Air Blank	0.000	01:17		
Diagnostics Check	OK	01:17		

Data Download Date/Time: 03/17/2016 16:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,SHAWN,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DICARLO,JOSEPH
ARRESTING OFFICER AGENCY: Winter Park PD
LAST AGENCY INSPECTION: 02/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: GAGNE

DATE/TIME OF TEST: 03/12/2016 01:50:37

Diagnostics Check	OK	01:43
Air Blank	0.000	01:43
Control Test	0.081	01:43
Air Blank	0.000	01:44
Subject Sample # 1	0.254	01:44
Breath Volume	1.464	01:44
Air Blank	0.000	01:45
Air Blank	0.000	01:47
Subject Sample # 2	0.256	01:48
Breath Volume	1.109	01:48
Air Blank	0.000	01:49
Control Test	0.081	01:49
Air Blank	0.000	01:50
Diagnostics Check	OK	01:50

Data Download Date/Time: 03/17/2016 16:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITFIELD, TIARA, D
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: WILSON, RUDOLPH
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 02/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:18

SUBJECT'S LAST NAME: VERSCH

DATE/TIME OF TEST: 03/12/2016 03:49:05

Diagnostics Check	OK	03:42
Air Blank	0.000	03:42
Control Test	0.083	03:43
Air Blank	0.000	03:43
Subject Sample # 1	0.000	03:44
Breath Volume	1.480	03:44
Air Blank	0.000	03:44
Air Blank	0.000	03:46
Subject Sample # 2	0.000	03:47
Breath Volume	1.730	03:47
Air Blank	0.000	03:48
Control Test	0.082	03:48
Air Blank	0.000	03:48
Diagnostics Check	OK	03:49

Data Download Date/Time: 03/17/2016 16:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,SHAWN,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: EDDY,N
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 02/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:53

SUBJECT'S LAST NAME: LANDIN

DATE/TIME OF TEST: 03/12/2016 04:34:45

Diagnostics Check	OK	04:23	
Air Blank	0.000	04:23	
Control Test	0.082	04:24	
Air Blank	0.000	04:24	
Subject Sample	# 1	04:27	No Sample Provided
Breath Volume	0.000	04:27	
Air Blank	0.000	04:28	
Air Blank	0.000	04:30	
Subject Sample	# 2	04:33	No Sample Provided
Breath Volume	0.000	04:33	
Air Blank	0.000	04:33	
Control Test	0.083	04:34	
Air Blank	0.000	04:34	
Diagnostics Check	OK	04:34	

Data Download Date/Time: 03/17/2016 16:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITFIELD, TIARA, D
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: LAL, RASIV
ARRESTING OFFICER AGENCY: Winter Garden PD
LAST AGENCY INSPECTION: 02/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:12

SUBJECT'S LAST NAME: IRWIN

DATE/TIME OF TEST: 03/13/2016 04:42:09

Diagnostics Check	OK	04:35
Air Blank	0.000	04:36
Control Test	0.082	04:36
Air Blank	0.000	04:37
Subject Sample # 1	0.186	04:37
Breath Volume	1.460	04:37
Air Blank	0.000	04:38
Air Blank	0.000	04:40
Subject Sample # 2	0.185	04:40
Breath Volume	1.152	04:40
Air Blank	0.000	04:41
Control Test	0.081	04:41
Air Blank	0.000	04:42
Diagnostics Check	OK	04:42

Data Download Date/Time: 03/17/2016 16:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: LOPEZ, JUSTINA
ARRESTING OFFICER AGENCY: Winter Park PD
LAST AGENCY INSPECTION: 02/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:52

SUBJECT'S LAST NAME: OGLE

DATE/TIME OF TEST: 03/13/2016 05:18:55

Diagnostics Check		OK	05:12
Air Blank		0.000	05:13
Control Test		0.083	05:13
Air Blank		0.000	05:13
Subject Sample	# 1	0.144	05:14
Breath Volume		1.804	05:14
Air Blank		0.000	05:15
Air Blank		0.000	05:16
Subject Sample	# 2	0.143	05:17
Breath Volume		2.980	05:17
Air Blank		0.000	05:18
Control Test		0.081	05:18
Air Blank		0.000	05:18
Diagnostics Check		OK	05:18

Data Download Date/Time: 03/17/2016 16:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FIGUEROA, HECTOR, A
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HAWKINS, DEBORAH
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 02/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 06:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:07

SUBJECT'S LAST NAME: GOMEZNAVARRETE

DATE/TIME OF TEST: 03/13/2016 07:37:17

Diagnostics Check	OK	07:30
Air Blank	0.000	07:31
Control Test	0.082	07:31
Air Blank	0.000	07:31
Subject Sample # 1	0.157	07:32
Breath Volume	1.375	07:32
Air Blank	0.000	07:33
Air Blank	0.000	07:35
Subject Sample # 2	0.163	07:35
Breath Volume	1.660	07:35
Air Blank	0.000	07:36
Control Test	0.082	07:36
Air Blank	0.000	07:37
Diagnostics Check	OK	07:37

Data Download Date/Time: 03/17/2016 16:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLEMAN,BRETT,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BALZAN,JEREMY
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 02/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:02

SUBJECT'S LAST NAME: HAMDAN

DATE/TIME OF TEST: 03/14/2016 04:28:51

Diagnostics Check	OK	04:22
Air Blank	0.000	04:22
Control Test	0.083	04:23
Air Blank	0.000	04:23
Subject Sample # 1	0.178	04:24
Breath Volume	1.648	04:24
Air Blank	0.000	04:25
Air Blank	0.000	04:26
Subject Sample # 2	0.184	04:27
Breath Volume	1.417	04:27
Air Blank	0.000	04:27
Control Test	0.081	04:28
Air Blank	0.000	04:28
Diagnostics Check	OK	04:28

Data Download Date/Time: 03/17/2016 16:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITFIELD, TIARA, D
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: GORBERG, JASON
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 02/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:17

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 03/16/2016 02:59:15

Diagnostics Check		OK	02:47	
Air Blank		0.000	02:48	
Control Test		0.082	02:48	
Air Blank		0.000	02:48	
Subject Sample	# 1	0.162	02:52	Volume Not Met
Breath Volume		0.000	02:52	
Air Blank		0.000	02:52	
Air Blank		0.000	02:54	
Subject Sample	# 2	0.167	02:57	Volume Not Met
Breath Volume		0.226	02:57	
Air Blank		0.000	02:58	
Control Test		0.082	02:58	
Air Blank		0.000	02:59	
Diagnostics Check		OK	02:59	

Data Download Date/Time: 03/17/2016 16:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FIGUEROA, HECTOR, A
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: KEANE, PETER
ARRESTING OFFICER AGENCY: Winter Garden PD
LAST AGENCY INSPECTION: 02/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:29

SUBJECT'S LAST NAME: SANTIAGO

DATE/TIME OF TEST: 03/16/2016 04:57:57

Diagnostics Check		OK	04:51
Air Blank		0.000	04:52
Control Test		0.082	04:52
Air Blank		0.000	04:53
Subject Sample	# 1	0.201	04:53
Breath Volume		2.414	04:53
Air Blank		0.000	04:54
Air Blank		0.000	04:55
Subject Sample	# 2	0.192	04:56
Breath Volume		2.792	04:56
Air Blank		0.000	04:57
Control Test		0.082	04:57
Air Blank		0.000	04:57
Diagnostics Check		OK	04:57

Data Download Date/Time: 03/17/2016 16:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HALLOCK, STEPHEN
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 02/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:11

SUBJECT'S LAST NAME: BOSTON

DATE/TIME OF TEST: 03/17/2016 00:38:59

Diagnostics Check	OK	00:32
Air Blank	0.000	00:33
Control Test	0.082	00:33
Air Blank	0.000	00:33
Subject Sample # 1	0.162	00:34
Breath Volume	1.300	00:34
Air Blank	0.000	00:35
Air Blank	0.000	00:36
Subject Sample # 2	0.161	00:37
Breath Volume	3.000	00:37
Air Blank	0.000	00:38
Control Test	0.081	00:38
Air Blank	0.000	00:38
Diagnostics Check	OK	00:38

Data Download Date/Time: 03/17/2016 16:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITFIELD, TIARA, D
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: GORBERG, JASON
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 02/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: HILL

DATE/TIME OF TEST: 03/17/2016 01:20:10

Diagnostics Check	OK	01:13
Air Blank	0.000	01:14
Control Test	0.082	01:14
Air Blank	0.000	01:15
Subject Sample # 1	0.072	01:15
Breath Volume	1.773	01:15
Air Blank	0.000	01:16
Air Blank	0.000	01:18
Subject Sample # 2	0.073	01:18
Breath Volume	3.207	01:18
Air Blank	0.000	01:19
Control Test	0.082	01:19
Air Blank	0.000	01:20
Diagnostics Check	OK	01:20

Data Download Date/Time: 03/17/2016 16:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITFIELD, TIARA, D
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: KIPP, ALEXANDER
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 02/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:23

SUBJECT'S LAST NAME: RALDA

DATE/TIME OF TEST: 03/18/2016 01:53:13

Diagnostics Check	OK	01:46
Air Blank	0.000	01:47
Control Test	0.083	01:47
Air Blank	0.000	01:47
Subject Sample # 1	0.159	01:48
Breath Volume	1.531	01:48
Air Blank	0.000	01:49
Air Blank	0.000	01:51
Subject Sample # 2	0.160	01:51
Breath Volume	1.511	01:51
Air Blank	0.000	01:52
Control Test	0.082	01:52
Air Blank	0.000	01:53
Diagnostics Check	OK	01:53

Data Download Date/Time: 03/22/2016 10:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HUTTO, GERALD
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 02/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: COLON

DATE/TIME OF TEST: 03/18/2016 02:17:02

Diagnostics Check	OK	02:10
Air Blank	0.000	02:10
Control Test	0.081	02:10
Air Blank	0.000	02:11
Subject Sample # 1	0.160	02:12
Breath Volume	1.687	02:12
Air Blank	0.000	02:12
Air Blank	0.000	02:14
Subject Sample # 2	0.157	02:15
Breath Volume	1.812	02:15
Air Blank	0.000	02:16
Control Test	0.081	02:16
Air Blank	0.000	02:16
Diagnostics Check	OK	02:17

Data Download Date/Time: 03/22/2016 10:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITFIELD, TIARA, D
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: KRALICH, BRENDEN
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 02/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: LEWIS

DATE/TIME OF TEST: 03/18/2016 03:38:20

Diagnostics Check	OK	03:26	
Air Blank	0.000	03:27	
Control Test	0.080	03:27	
Air Blank	0.000	03:28	
Subject Sample	# 1	03:31	No Sample Provided
Breath Volume	0.000	03:31	
Air Blank	0.000	03:31	
Air Blank	0.000	03:33	
Subject Sample	# 2	03:36	No Sample Provided
Breath Volume	0.000	03:36	
Air Blank	0.000	03:37	
Control Test	0.082	03:37	
Air Blank	0.000	03:38	
Diagnostics Check	OK	03:38	

Data Download Date/Time: 03/22/2016 10:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: RODRIGUEZ, GEORGE
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 02/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: ROBILLARD

DATE/TIME OF TEST: 03/18/2016 04:19:08

Diagnostics Check	OK	04:07	
Air Blank	0.000	04:08	
Control Test	0.081	04:08	
Air Blank	0.000	04:09	
Subject Sample	# 1	04:12	No Sample Provided
Breath Volume	0.000	04:12	
Air Blank	0.000	04:12	
Air Blank	0.000	04:14	
Subject Sample	# 2	04:17	No Sample Provided
Breath Volume	0.000	04:17	
Air Blank	0.000	04:18	
Control Test	0.082	04:18	
Air Blank	0.000	04:19	
Diagnostics Check	OK	04:19	

Data Download Date/Time: 03/22/2016 10:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITFIELD, TIARA, D
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MOONDA, SHUAIB
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 02/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:52

SUBJECT'S LAST NAME: NYUNT

DATE/TIME OF TEST: 03/18/2016 05:22:46

Diagnostics Check	OK	05:16
Air Blank	0.000	05:16
Control Test	0.082	05:16
Air Blank	0.000	05:17
Subject Sample # 1	0.169	05:18
Breath Volume	2.136	05:18
Air Blank	0.000	05:18
Air Blank	0.000	05:20
Subject Sample # 2	0.173	05:21
Breath Volume	2.667	05:21
Air Blank	0.000	05:21
Control Test	0.080	05:22
Air Blank	0.000	05:22
Diagnostics Check	OK	05:22

Data Download Date/Time: 03/22/2016 10:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: PINO PEREZ, ALEJANDRO
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 02/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:58

SUBJECT'S LAST NAME: ENRIQUE ALEJANDRO

DATE/TIME OF TEST: 03/19/2016 01:27:58

Diagnostics Check	OK	01:21
Air Blank	0.000	01:22
Control Test	0.082	01:22
Air Blank	0.000	01:23
Subject Sample # 1	0.078	01:23
Breath Volume	1.261	01:23
Air Blank	0.000	01:24
Air Blank	0.000	01:26
Subject Sample # 2	0.076	01:26
Breath Volume	1.621	01:26
Air Blank	0.000	01:27
Control Test	0.082	01:27
Air Blank	0.000	01:27
Diagnostics Check	OK	01:27

Data Download Date/Time: 03/22/2016 10:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: URIBE, SERGIO
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 02/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:25

SUBJECT'S LAST NAME: GANDY

DATE/TIME OF TEST: 03/19/2016 04:59:59

Diagnostics Check		OK	04:48	
Air Blank		0.000	04:48	
Control Test		0.082	04:49	
Air Blank		0.000	04:49	
Subject Sample	# 1	0.068	04:52	Volume Not Met
Breath Volume		0.097	04:52	
Air Blank		0.000	04:53	
Air Blank		0.000	04:55	
Subject Sample	# 2	0.071	04:58	Volume Not Met
Breath Volume		0.210	04:58	
Air Blank		0.000	04:59	
Control Test		0.080	04:59	
Air Blank		0.000	04:59	
Diagnostics Check		OK	04:59	

Data Download Date/Time: 03/22/2016 10:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: VAUGHN, JACOB
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 02/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: VERGARA

DATE/TIME OF TEST: 03/20/2016 00:08:23

Diagnostics Check	OK	00:05	
Air Blank	0.000	00:05	
Control Test	0.082	00:06	
Air Blank	0.000	00:06	
Subject Sample	# 1	00:07	Subject Test Refused
Breath Volume	0.000	00:07	
Air Blank	0.000	00:07	
Control Test	0.082	00:07	
Air Blank	0.000	00:08	
Diagnostics Check	OK	00:08	

Data Download Date/Time: 03/22/2016 10:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FIGUEROA, HECTOR, A
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BARRIENTOS, EVELYN
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 02/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: MOBEEN

DATE/TIME OF TEST: 03/20/2016 01:17:59

Diagnostics Check	OK	01:10
Air Blank	0.000	01:11
Control Test	0.082	01:11
Air Blank	0.000	01:11
Subject Sample # 1	0.084	01:13
Breath Volume	1.226	01:13
Air Blank	0.000	01:13
Air Blank	0.000	01:15
Subject Sample # 2	0.085	01:16
Breath Volume	1.562	01:16
Air Blank	0.000	01:17
Control Test	0.081	01:17
Air Blank	0.000	01:17
Diagnostics Check	OK	01:17

Data Download Date/Time: 03/22/2016 10:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FIGUEROA, HECTOR, A
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: COOP, JOHN
ARRESTING OFFICER AGENCY: Winter Park PD
LAST AGENCY INSPECTION: 02/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:00

SUBJECT'S LAST NAME: EXANTUS

DATE/TIME OF TEST: 03/20/2016 05:36:43

Diagnostics Check	OK	05:28
Air Blank	0.000	05:29
Control Test	0.082	05:29
Air Blank	0.000	05:29
Subject Sample # 1	0.205	05:30
Breath Volume	1.285	05:30
Air Blank	0.000	05:31
Air Blank	0.000	05:32
Subject Sample # 2	0.204	05:35
Breath Volume	1.468	05:35
Air Blank	0.000	05:35
Control Test	0.082	05:36
Air Blank	0.000	05:36
Diagnostics Check	OK	05:36

Data Download Date/Time: 03/22/2016 10:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITFIELD, TIARA, D
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ANDREWS, GARY
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 02/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:38

SUBJECT'S LAST NAME: ROBINSON

DATE/TIME OF TEST: 03/22/2016 03:14:00

Diagnostics Check		OK	03:02	
Air Blank		0.000	03:02	
Control Test		0.083	03:03	
Air Blank		0.000	03:03	
Subject Sample	# 1	0.125	03:06	Volume Not Met
Breath Volume		0.347	03:06	
Air Blank		0.000	03:07	
Air Blank		0.000	03:09	
Subject Sample	# 2	0.125	03:12	Volume Not Met
Breath Volume		0.148	03:12	
Air Blank		0.000	03:13	
Control Test		0.081	03:13	
Air Blank		0.000	03:13	
Diagnostics Check		OK	03:14	

Data Download Date/Time: 03/22/2016 10:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FIGUEROA, HECTOR, A
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: TURNER, MICHAEL
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 02/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: HOWARD

DATE/TIME OF TEST: 03/22/2016 03:33:56

Diagnostics Check	OK	03:27
Air Blank	0.000	03:28
Control Test	0.081	03:28
Air Blank	0.000	03:29
Subject Sample # 1	0.273	03:29
Breath Volume	3.582	03:29
Air Blank	0.000	03:30
Air Blank	0.000	03:31
Subject Sample # 2	0.265	03:32
Breath Volume	1.402	03:32
Air Blank	0.000	03:33
Control Test	0.081	03:33
Air Blank	0.000	03:33
Diagnostics Check	OK	03:33

Data Download Date/Time: 03/22/2016 10:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDEN,WILLIAM,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: PRICE,FOREST
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 03/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:19

SUBJECT'S LAST NAME: ROSA FLORES

DATE/TIME OF TEST: 03/23/2016 02:46:58

Diagnostics Check		OK	02:40
Air Blank		0.000	02:40
Control Test		0.082	02:41
Air Blank		0.000	02:41
Subject Sample	# 1	0.000	02:42
Breath Volume		2.007	02:42
Air Blank		0.000	02:42
Air Blank		0.000	02:44
Subject Sample	# 2	0.000	02:45
Breath Volume		1.117	02:45
Air Blank		0.000	02:46
Control Test		0.082	02:46
Air Blank		0.000	02:46
Diagnostics Check		OK	02:46

Data Download Date/Time: 03/31/2016 15:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDEN,WILLIAM,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BENET,MARLON
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 03/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: JEAN-GILLES

DATE/TIME OF TEST: 03/24/2016 01:55:52

Diagnostics Check	OK	01:49
Air Blank	0.000	01:49
Control Test	0.083	01:50
Air Blank	0.000	01:50
Subject Sample # 1	0.000	01:51
Breath Volume	2.851	01:51
Air Blank	0.000	01:51
Air Blank	0.000	01:53
Subject Sample # 2	0.000	01:54
Breath Volume	1.878	01:54
Air Blank	0.000	01:54
Control Test	0.082	01:55
Air Blank	0.000	01:55
Diagnostics Check	OK	01:55

Data Download Date/Time: 03/31/2016 15:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDEN,WILLIAM,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HAWKINS,DEBBIE
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 03/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:52

SUBJECT'S LAST NAME: LEMUS-ALVAREZ

DATE/TIME OF TEST: 03/24/2016 23:19:16

Diagnostics Check	OK	23:12
Air Blank	0.000	23:13
Control Test	0.082	23:13
Air Blank	0.000	23:14
Subject Sample # 1	0.279	23:14
Breath Volume	3.210	23:14
Air Blank	0.000	23:15
Air Blank	0.000	23:17
Subject Sample # 2	0.278	23:17
Breath Volume	3.160	23:17
Air Blank	0.000	23:18
Control Test	0.081	23:18
Air Blank	0.000	23:19
Diagnostics Check	OK	23:19

Data Download Date/Time: 03/31/2016 15:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDEN,WILLIAM,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: WEAVER,WILLIAM
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 03/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:56

SUBJECT'S LAST NAME: BRANSCOMB

DATE/TIME OF TEST: 03/26/2016 01:24:36

Diagnostics Check	OK	01:18
Air Blank	0.000	01:18
Control Test	0.082	01:19
Air Blank	0.000	01:19
Subject Sample # 1	0.000	01:20
Breath Volume	1.937	01:20
Air Blank	0.000	01:20
Air Blank	0.000	01:22
Subject Sample # 2	0.000	01:23
Breath Volume	2.175	01:23
Air Blank	0.000	01:23
Control Test	0.082	01:24
Air Blank	0.000	01:24
Diagnostics Check	OK	01:24

Data Download Date/Time: 03/31/2016 15:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,SHAWN,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: IMPARATO,FRANK
ARRESTING OFFICER AGENCY: Central FL Univ PD
LAST AGENCY INSPECTION: 03/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:52

SUBJECT'S LAST NAME: ARBELAEZ MEDINA

DATE/TIME OF TEST: 03/26/2016 02:28:02

Diagnostics Check		OK	02:16	
Air Blank		0.000	02:16	
Control Test		0.081	02:17	
Air Blank		0.000	02:17	
Subject Sample	# 1	0.153	02:20	Volume Not Met
Breath Volume		0.097	02:20	
Air Blank		0.000	02:21	
Air Blank		0.000	02:23	
Subject Sample	# 2	0.141	02:26	Volume Not Met
Breath Volume		0.066	02:26	
Air Blank		0.000	02:27	
Control Test		0.079	02:27	
Air Blank		0.000	02:27	
Diagnostics Check		OK	02:28	

Data Download Date/Time: 03/31/2016 15:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDEN,WILLIAM,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BINET,MARLON
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 03/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:26

SUBJECT'S LAST NAME: BASSAN

DATE/TIME OF TEST: 03/26/2016 02:53:04

Diagnostics Check	OK	02:48	
Air Blank	0.000	02:48	
Control Test	0.081	02:48	
Air Blank	0.000	02:49	
Subject Sample	# 1	02:51	Subject Test Refused
Breath Volume	0.000	02:51	
Air Blank	0.000	02:52	
Control Test	0.081	02:52	
Air Blank	0.000	02:52	
Diagnostics Check	OK	02:53	

Data Download Date/Time: 03/31/2016 15:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,SHAWN,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DOBBINS,SEAN
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 03/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:26

SUBJECT'S LAST NAME: AUDIOIRE

DATE/TIME OF TEST: 03/26/2016 03:55:22

Diagnostics Check	OK	03:49
Air Blank	0.000	03:49
Control Test	0.080	03:49
Air Blank	0.000	03:50
Subject Sample # 1	0.260	03:50
Breath Volume	3.714	03:50
Air Blank	0.000	03:51
Air Blank	0.000	03:53
Subject Sample # 2	0.257	03:53
Breath Volume	2.753	03:53
Air Blank	0.000	03:54
Control Test	0.079	03:54
Air Blank	0.000	03:55
Diagnostics Check	OK	03:55

Data Download Date/Time: 03/31/2016 15:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ASHE,KEVIN,B
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: PRITCHARD,T
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 02/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 19:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: BUCHANAN

DATE/TIME OF TEST: 03/02/2016 20:26:33

Diagnostics Check	OK	20:23	
Air Blank	0.000	20:23	
Control Test	0.081	20:24	
Air Blank	0.000	20:24	
Subject Sample	# 1	20:25	Subject Test Refused
Breath Volume	0.000	20:25	
Air Blank	0.000	20:25	
Control Test	0.080	20:25	
Air Blank	0.000	20:26	
Diagnostics Check	OK	20:26	

Data Download Date/Time: 03/21/2016 16:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ASHE,KEVIN,B
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: SWIM,GREG
ARRESTING OFFICER AGENCY: Orange Park PD
LAST AGENCY INSPECTION: 02/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: CARTER

DATE/TIME OF TEST: 03/03/2016 02:27:53

Diagnostics Check	OK	02:16	
Air Blank	0.000	02:16	
Control Test	0.082	02:17	
Air Blank	0.000	02:17	
Subject Sample	# 1	02:20	No Sample Provided
Breath Volume	0.000	02:20	
Air Blank	0.000	02:21	
Air Blank	0.000	02:23	
Subject Sample	# 2	02:26	No Sample Provided
Breath Volume	0.000	02:26	
Air Blank	0.000	02:26	
Control Test	0.082	02:27	
Air Blank	0.000	02:27	
Diagnostics Check	OK	02:27	

Data Download Date/Time: 03/21/2016 16:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALTSTATT, MATTHEW, B
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: LOVETT, T
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 02/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:49

SUBJECT'S LAST NAME: LASTER

DATE/TIME OF TEST: 03/11/2016 04:17:06

Diagnostics Check	OK	04:09
Air Blank	0.000	04:10
Control Test	0.081	04:10
Air Blank	0.000	04:11
Subject Sample # 1	0.213	04:12
Breath Volume	2.082	04:12
Air Blank	0.000	04:12
Air Blank	0.000	04:14
Subject Sample # 2	0.214	04:15
Breath Volume	1.910	04:15
Air Blank	0.000	04:16
Control Test	0.081	04:16
Air Blank	0.000	04:17
Diagnostics Check	OK	04:17

Data Download Date/Time: 03/21/2016 16:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RANDOLPH,K,J
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: MASON,M
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 02/21/2016

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 13:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:04

SUBJECT'S LAST NAME: PASCHAL

DATE/TIME OF TEST: 03/15/2016 14:34:23

Diagnostics Check	OK	14:31	
Air Blank	0.000	14:31	
Control Test	0.081	14:31	
Air Blank	0.000	14:32	
Subject Sample # 1	0.257	14:32	
Breath Volume	2.253	14:32	
Air Blank		14:32	Improper Sample
Air Blank	0.000	14:33	
Control Test	0.081	14:33	
Air Blank	0.000	14:34	
Diagnostics Check	OK	14:34	

Data Download Date/Time: 03/21/2016 16:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ASHE,KEVIN,B
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: PRITCHARD,T
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 03/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: ADAMS

DATE/TIME OF TEST: 03/22/2016 02:42:01

Diagnostics Check	OK	02:35
Air Blank	0.000	02:36
Control Test	0.081	02:36
Air Blank	0.000	02:36
Subject Sample # 1	0.103	02:37
Breath Volume	3.679	02:37
Air Blank	0.000	02:37
Air Blank	0.000	02:39
Subject Sample # 2	0.100	02:40
Breath Volume	3.222	02:40
Air Blank	0.000	02:41
Control Test	0.081	02:41
Air Blank	0.000	02:41
Diagnostics Check	OK	02:42

Data Download Date/Time: 04/27/2016 14:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ASHE,KEVIN,B
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: GLENN,CHRIS
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 03/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: PLOURDE

DATE/TIME OF TEST: 03/27/2016 03:32:27

Diagnostics Check	OK	03:21	
Air Blank	0.000	03:21	
Control Test	0.080	03:21	
Air Blank	0.000	03:22	
Subject Sample	# 1	03:25	No Sample Provided
Breath Volume	0.000	03:25	
Air Blank	0.000	03:25	
Air Blank	0.000	03:27	
Subject Sample	# 2	03:31	No Sample Provided
Breath Volume	0.000	03:31	
Air Blank	0.000	03:31	
Control Test	0.080	03:31	
Air Blank	0.000	03:32	
Diagnostics Check	OK	03:32	

Data Download Date/Time: 04/27/2016 14:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,JOHN,K
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: SAUNDERS,J
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 03/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 19:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:15

SUBJECT'S LAST NAME: MORGAN

DATE/TIME OF TEST: 03/28/2016 20:49:09

Diagnostics Check	OK	20:42
Air Blank	0.000	20:42
Control Test	0.080	20:43
Air Blank	0.000	20:43
Subject Sample # 1	0.000	20:44
Breath Volume	1.121	20:44
Air Blank	0.000	20:44
Air Blank	0.000	20:46
Subject Sample # 2	0.000	20:47
Breath Volume	1.542	20:47
Air Blank	0.000	20:48
Control Test	0.079	20:48
Air Blank	0.000	20:49
Diagnostics Check	OK	20:49

Data Download Date/Time: 04/27/2016 14:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ASHE,KEVIN,B
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: PRITCHARD,T
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 03/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 19:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:53

SUBJECT'S LAST NAME: LONER

DATE/TIME OF TEST: 03/31/2016 20:21:17

Diagnostics Check		OK	20:14
Air Blank		0.000	20:15
Control Test		0.081	20:15
Air Blank		0.000	20:16
Subject Sample	# 1	0.248	20:16
Breath Volume		2.132	20:16
Air Blank		0.000	20:17
Air Blank		0.000	20:19
Subject Sample	# 2	0.250	20:19
Breath Volume		2.328	20:19
Air Blank		0.000	20:20
Control Test		0.080	20:20
Air Blank		0.000	20:21
Diagnostics Check		OK	20:21

Data Download Date/Time: 04/27/2016 14:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001646

INSTRUMENT REGISTERED TO: TYNDALL AFB
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUBER,PHILLIP,W
BREATH TEST OPERATOR AGENCY: FDLE
ARRESTING OFFICER: SUBER,PHILLIP
ARRESTING OFFICER AGENCY: FDLE
LAST AGENCY INSPECTION: 02/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A2
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: SYS CHECK
ARREST TIME: 08:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:40

SUBJECT'S LAST NAME: LICENSE

DATE/TIME OF TEST: 03/10/2016 09:07:53

Diagnostics Check	OK	09:01
Air Blank	0.000	09:02
Control Test	0.078	09:02
Air Blank	0.000	09:03
Subject Sample # 1	0.000	09:03
Breath Volume	1.542	09:03
Air Blank	0.000	09:04
Air Blank	0.000	09:06
Subject Sample # 2	0.000	09:06
Breath Volume	1.589	09:06
Air Blank	0.000	09:06
Control Test	0.078	09:07
Air Blank	0.000	09:07
Diagnostics Check	OK	09:07

Data Download Date/Time: 03/25/2016 13:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001646

INSTRUMENT REGISTERED TO: TYNDALL AFB
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUBER,PHILLIP,W
BREATH TEST OPERATOR AGENCY: FDLE
ARRESTING OFFICER: SUBER,PHILLIP
ARRESTING OFFICER AGENCY: FDLE
LAST AGENCY INSPECTION: 02/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: SYS CHECK
ARREST TIME: 13:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:50

SUBJECT'S LAST NAME: TEST

DATE/TIME OF TEST: 03/28/2016 14:24:26

Diagnostics Check		OK	14:18
Air Blank		0.000	14:19
Control Test		0.078	14:19
Air Blank		0.000	14:19
Subject Sample	# 1	0.000	14:20
Breath Volume		1.300	14:20
Air Blank		0.000	14:20
Air Blank		0.000	14:22
Subject Sample	# 2	0.000	14:23
Breath Volume		1.210	14:23
Air Blank		0.000	14:23
Control Test		0.078	14:23
Air Blank		0.000	14:24
Diagnostics Check		OK	14:24

Data Download Date/Time: 03/28/2016 17:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001654

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,ZANE,A
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: JONES,ZANE
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 02/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 23:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:21

SUBJECT'S LAST NAME: SILVERIO

DATE/TIME OF TEST: 03/08/2016 00:47:00

Diagnostics Check	OK	00:44	
Air Blank	0.000	00:44	
Control Test	0.082	00:44	
Air Blank	0.000	00:45	
Subject Sample	# 1	00:45	Subject Test Refused
Breath Volume	0.000	00:45	
Air Blank	0.000	00:46	
Control Test	0.081	00:46	
Air Blank	0.000	00:46	
Diagnostics Check	OK	00:47	

Data Download Date/Time: 03/10/2016 17:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001654

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,ZANE,A
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: JONES,ZANE
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 02/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 23:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:05

SUBJECT'S LAST NAME: CABRISAS

DATE/TIME OF TEST: 03/08/2016 23:43:17

Diagnostics Check		OK	23:31	
Air Blank		0.000	23:31	
Control Test		0.082	23:32	
Air Blank		0.000	23:32	
Subject Sample	# 1	0.132	23:35	Volume Not Met
Breath Volume		0.312	23:35	
Air Blank		0.000	23:36	
Air Blank		0.000	23:38	
Subject Sample	# 2	0.143	23:38	
Breath Volume		1.285	23:38	
Air Blank		0.000	23:39	
Air Blank		0.000	23:41	
Subject Sample	# 3	0.139	23:41	
Breath Volume		1.386	23:41	
Air Blank		0.000	23:42	
Control Test		0.079	23:42	
Air Blank		0.000	23:43	
Diagnostics Check		OK	23:43	

Data Download Date/Time: 03/10/2016 17:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001654

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,ZANE,A
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: JONES,ZANE
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 03/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 23:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: EUGENE

DATE/TIME OF TEST: 03/10/2016 23:59:26

Diagnostics Check	OK	23:53	
Air Blank	0.000	23:53	
Control Test	0.082	23:54	
Air Blank	0.000	23:54	
Subject Sample # 1	0.144	23:55	
Breath Volume	1.144	23:55	
Air Blank	0.000	23:56	
Air Blank		23:57	Improper Sample
Air Blank	0.000	23:58	
Control Test	0.082	23:58	
Air Blank	0.000	23:59	
Diagnostics Check	OK	23:59	

Data Download Date/Time: 04/22/2016 18:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001654

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,ZANE,A
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: JONES,ZANE
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 03/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 23:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: EUGENE

DATE/TIME OF TEST: 03/11/2016 00:09:40

Diagnostics Check	OK	00:02
Air Blank	0.000	00:02
Control Test	0.080	00:02
Air Blank	0.000	00:03
Subject Sample # 1	0.160	00:04
Breath Volume	2.171	00:04
Air Blank	0.000	00:05
Air Blank	0.000	00:06
Subject Sample # 2	0.163	00:07
Breath Volume	3.242	00:07
Air Blank	0.000	00:08
Control Test	0.079	00:09
Air Blank	0.000	00:09
Diagnostics Check	OK	00:09

Data Download Date/Time: 04/22/2016 18:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001654

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ASCANIO, MARCEL, S
BREATH TEST OPERATOR AGENCY: North Bay Village PD
ARRESTING OFFICER: ASCANIO, MARCEL
ARRESTING OFFICER AGENCY: North Bay Village PD
LAST AGENCY INSPECTION: 03/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 06:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:56

SUBJECT'S LAST NAME: ALIAGA ARAUJO

DATE/TIME OF TEST: 03/13/2016 07:30:37

Diagnostics Check		OK	07:17	
Air Blank		0.000	07:18	
Control Test		0.081	07:18	
Air Blank		0.000	07:18	
Subject Sample	# 1	0.214	07:22	Volume Not Met
Breath Volume		0.035	07:22	
Air Blank		0.000	07:22	
Air Blank		0.000	07:24	
Subject Sample	# 2	0.238	07:25	
Breath Volume		1.800	07:25	
Air Blank		0.000	07:26	
Air Blank		0.000	07:28	
Subject Sample	# 3	0.254	07:29	
Breath Volume		1.589	07:29	
Air Blank		0.000	07:29	
Control Test		0.079	07:30	
Air Blank		0.000	07:30	
Diagnostics Check		OK	07:30	

Data Download Date/Time: 04/22/2016 18:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001654

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: REFUSE,LUDDWIDGE,

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: CARRIEL,RANDY

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 03/10/2016

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1789170

DRY GAS EXPIRY DATE: 01/14/2018

VIOLATION CODE: DUI

ARREST TIME: 23:31

GENDER: Male

OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: CELESTIN

DATE/TIME OF TEST: 03/27/2016 00:37:02

Diagnostics Check	OK	00:30	
Air Blank	0.000	00:31	
Control Test	0.081	00:31	
Air Blank	0.000	00:31	
Subject Sample	# 1 0.124	00:32	
Breath Volume	1.582	00:32	
Air Blank	0.000	00:33	
Air Blank	0.000	00:34	
Subject Sample	# 2 0.130	00:36	
Breath Volume	2.269	00:36	
Air Blank		00:36	RFI Detect
Air Blank	0.000	00:37	

Data Download Date/Time: 04/22/2016 18:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASH,JUSTIN,A
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: GOLDMAN,JEREMY
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 02/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1568991
DRY GAS EXPIRY DATE: 10/01/2016
VIOLATION CODE: DUI
ARREST TIME: 19:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:15

SUBJECT'S LAST NAME: GULYAS

DATE/TIME OF TEST: 03/02/2016 20:43:37

Diagnostics Check	OK	20:37
Air Blank	0.000	20:37
Control Test	0.079	20:38
Air Blank	0.000	20:38
Subject Sample # 1	0.197	20:39
Breath Volume	1.585	20:39
Air Blank	0.000	20:39
Air Blank	0.000	20:41
Subject Sample # 2	0.192	20:42
Breath Volume	2.054	20:42
Air Blank	0.000	20:42
Control Test	0.078	20:43
Air Blank	0.000	20:43
Diagnostics Check	OK	20:43

Data Download Date/Time: 03/31/2016 13:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,CHRIS,
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: MILLER,CHRIS
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 02/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1568991
DRY GAS EXPIRY DATE: 10/01/2016
VIOLATION CODE: DUI
ARREST TIME: 07:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:01

SUBJECT'S LAST NAME: LAKE

DATE/TIME OF TEST: 03/10/2016 08:28:38

Diagnostics Check	OK	08:22
Air Blank	0.000	08:22
Control Test	0.080	08:23
Air Blank	0.000	08:23
Subject Sample # 1	0.165	08:23
Breath Volume	1.578	08:23
Air Blank	0.000	08:24
Air Blank	0.000	08:26
Subject Sample # 2	0.163	08:27
Breath Volume	1.324	08:27
Air Blank	0.000	08:27
Control Test	0.081	08:28
Air Blank	0.000	08:28
Diagnostics Check	OK	08:28

Data Download Date/Time: 03/31/2016 13:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,SCOTT,B
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MILLER,SCOTT
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 02/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1568991
DRY GAS EXPIRY DATE: 10/01/2016
VIOLATION CODE: DUI
ARREST TIME: 15:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:10

SUBJECT'S LAST NAME: CULLIPHER

DATE/TIME OF TEST: 03/10/2016 16:38:45

Diagnostics Check	OK	16:32
Air Blank	0.000	16:32
Control Test	0.081	16:33
Air Blank	0.000	16:33
Subject Sample # 1	0.338	16:34
Breath Volume	2.230	16:34
Air Blank	0.000	16:34
Air Blank	0.000	16:36
Subject Sample # 2	0.341	16:37
Breath Volume	1.765	16:37
Air Blank	0.000	16:37
Control Test	0.079	16:38
Air Blank	0.000	16:38
Diagnostics Check	OK	16:38

Data Download Date/Time: 03/31/2016 13:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSON,GARETH,E
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: FISCHER,STEVE
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 02/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1568991
DRY GAS EXPIRY DATE: 10/01/2016
VIOLATION CODE: DUI
ARREST TIME: 19:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:40

SUBJECT'S LAST NAME: MORAN

DATE/TIME OF TEST: 03/11/2016 21:06:55

Diagnostics Check	OK	21:00
Air Blank	0.000	21:01
Control Test	0.080	21:01
Air Blank	0.000	21:02
Subject Sample # 1	0.203	21:02
Breath Volume	1.992	21:02
Air Blank	0.000	21:03
Air Blank	0.000	21:05
Subject Sample # 2	0.187	21:05
Breath Volume	1.847	21:05
Air Blank	0.000	21:06
Control Test	0.079	21:06
Air Blank	0.000	21:06
Diagnostics Check	OK	21:06

Data Download Date/Time: 03/31/2016 13:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSON,GARETH,E
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: JOHNSON,GARETH
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 02/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1568991
DRY GAS EXPIRY DATE: 10/01/2016
VIOLATION CODE: DUI
ARREST TIME: 02:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: WADE

DATE/TIME OF TEST: 03/12/2016 03:08:46

Diagnostics Check	OK	03:02
Air Blank	0.000	03:02
Control Test	0.081	03:03
Air Blank	0.000	03:03
Subject Sample # 1	0.173	03:04
Breath Volume	1.656	03:04
Air Blank	0.000	03:04
Air Blank	0.000	03:06
Subject Sample # 2	0.156	03:07
Breath Volume	3.453	03:07
Air Blank	0.000	03:07
Control Test	0.079	03:08
Air Blank	0.000	03:08
Diagnostics Check	OK	03:08

Data Download Date/Time: 03/31/2016 13:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSON,GARETH,E
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: JOHNSON,GARETH
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 02/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1568991
DRY GAS EXPIRY DATE: 10/01/2016
VIOLATION CODE: DUI
ARREST TIME: 02:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:59

SUBJECT'S LAST NAME: ZIESEMER

DATE/TIME OF TEST: 03/18/2016 03:37:17

Diagnostics Check	OK	03:30
Air Blank	0.000	03:30
Control Test	0.080	03:30
Air Blank	0.000	03:31
Subject Sample # 1	0.199	03:32
Breath Volume	4.218	03:32
Air Blank	0.000	03:33
Air Blank	0.000	03:34
Subject Sample # 2	0.200	03:35
Breath Volume	3.144	03:35
Air Blank	0.000	03:36
Control Test	0.078	03:36
Air Blank	0.000	03:37
Diagnostics Check	OK	03:37

Data Download Date/Time: 03/31/2016 13:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KELLY,EZEKIEL,D
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: KELLY,EZEKIEL
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 02/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1568991
DRY GAS EXPIRY DATE: 10/01/2016
VIOLATION CODE: DUI
ARREST TIME: 20:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:36

SUBJECT'S LAST NAME: KEMMAN

DATE/TIME OF TEST: 03/28/2016 21:24:40

Diagnostics Check		OK	21:18
Air Blank		0.000	21:19
Control Test		0.075	21:19
Air Blank		0.000	21:19
Subject Sample	# 1	0.155	21:20
Breath Volume		2.226	21:20
Air Blank		0.000	21:20
Air Blank		0.000	21:22
Subject Sample	# 2	0.166	21:23
Breath Volume		2.226	21:23
Air Blank		0.000	21:23
Control Test		0.075	21:24
Air Blank		0.000	21:24
Diagnostics Check		OK	21:24

Data Download Date/Time: 03/31/2016 13:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LYONS,D,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: LYONS,D
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 02/08/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1589901
DRY GAS EXPIRY DATE: 10/30/2016
VIOLATION CODE: DUI
ARREST TIME: 04:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:21

SUBJECT'S LAST NAME: COLLINS

DATE/TIME OF TEST: 03/14/2016 04:49:11

Diagnostics Check	OK	04:41
Air Blank	0.000	04:42
Control Test	0.080	04:42
Air Blank	0.000	04:43
Subject Sample # 1	0.177	04:43
Breath Volume	1.554	04:43
Air Blank	0.000	04:44
Air Blank	0.000	04:46
Subject Sample # 2	0.170	04:47
Breath Volume	1.902	04:47
Air Blank	0.000	04:48
Control Test	0.079	04:48
Air Blank	0.000	04:49
Diagnostics Check	OK	04:49

Data Download Date/Time: 03/25/2016 16:10