

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000738

INSTRUMENT REGISTERED TO: FT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRAWFORD,BENJAMIN,D
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: CRAWFORD,BENJAMIN
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 02/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A1
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 12:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:50

SUBJECT'S LAST NAME: KENDALL

DATE/TIME OF TEST: 03/04/2016 13:14:22

Diagnostics Check	OK	13:11	
Air Blank	0.000	13:12	
Control Test	0.080	13:12	
Air Blank	0.000	13:12	
Subject Sample	# 1	13:13	Subject Test Refused
Breath Volume	0.000	13:13	
Air Blank	0.000	13:13	
Control Test	0.079	13:13	
Air Blank	0.000	13:14	
Diagnostics Check	OK	13:14	

Data Download Date/Time: 03/31/2016 22:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000738

INSTRUMENT REGISTERED TO: FT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESNEUL,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: WRIGHT,RENARD
ARRESTING OFFICER AGENCY: Hurlburt Field PD
LAST AGENCY INSPECTION: 02/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A1
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: BRYAN

DATE/TIME OF TEST: 03/05/2016 01:24:32

Diagnostics Check	OK	01:18
Air Blank	0.000	01:18
Control Test	0.081	01:19
Air Blank	0.000	01:19
Subject Sample # 1	0.179	01:20
Breath Volume	2.304	01:20
Air Blank	0.000	01:20
Air Blank	0.000	01:22
Subject Sample # 2	0.172	01:23
Breath Volume	1.773	01:23
Air Blank	0.000	01:23
Control Test	0.080	01:23
Air Blank	0.000	01:24
Diagnostics Check	OK	01:24

Data Download Date/Time: 03/31/2016 22:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000738

INSTRUMENT REGISTERED TO: FT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRUJILLO,ANTHONY,R
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: TRUJILLO,ANTHONY
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 02/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A1
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:25

SUBJECT'S LAST NAME: BELSON

DATE/TIME OF TEST: 03/05/2016 20:52:23

Diagnostics Check	OK	20:49	
Air Blank	0.000	20:50	
Control Test	0.081	20:50	
Air Blank	0.000	20:50	
Subject Sample	# 1	20:51	Subject Test Refused
Breath Volume	0.000	20:51	
Air Blank	0.000	20:51	
Control Test	0.079	20:51	
Air Blank	0.000	20:52	
Diagnostics Check	OK	20:52	

Data Download Date/Time: 03/31/2016 22:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000738

INSTRUMENT REGISTERED TO: FT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREEN,DONALD,H
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: GREEN,DONALD
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 02/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A1
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:07

SUBJECT'S LAST NAME: DUNN

DATE/TIME OF TEST: 03/08/2016 21:38:21

Diagnostics Check	OK	21:31
Air Blank	0.000	21:32
Control Test	0.080	21:32
Air Blank	0.000	21:33
Subject Sample # 1	0.000	21:34
Breath Volume	3.156	21:34
Air Blank	0.000	21:34
Air Blank	0.000	21:36
Subject Sample # 2	0.000	21:37
Breath Volume	2.722	21:37
Air Blank	0.000	21:37
Control Test	0.079	21:37
Air Blank	0.000	21:38
Diagnostics Check	OK	21:38

Data Download Date/Time: 03/31/2016 22:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000738

INSTRUMENT REGISTERED TO: FT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESNEUL,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: GONZALEZ,FREDERICO
ARRESTING OFFICER AGENCY: Hurlburt Field PD
LAST AGENCY INSPECTION: 02/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A1
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: BELDIN

DATE/TIME OF TEST: 03/12/2016 02:19:07

Diagnostics Check	OK	02:12
Air Blank	0.000	02:13
Control Test	0.080	02:13
Air Blank	0.000	02:14
Subject Sample # 1	0.124	02:14
Breath Volume	4.566	02:14
Air Blank	0.000	02:15
Air Blank	0.000	02:17
Subject Sample # 2	0.120	02:17
Breath Volume	3.718	02:17
Air Blank	0.000	02:18
Control Test	0.080	02:18
Air Blank	0.000	02:19
Diagnostics Check	OK	02:19

Data Download Date/Time: 03/31/2016 22:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000738

INSTRUMENT REGISTERED TO: FT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRUJILLO,ANTHONY,R
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: TRUJILLO,ANTHONY
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 02/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A1
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: BARGER

DATE/TIME OF TEST: 03/13/2016 04:05:30

Diagnostics Check	OK	04:02	
Air Blank	0.000	04:03	
Control Test	0.080	04:03	
Air Blank	0.000	04:03	
Subject Sample	# 1	04:04	Subject Test Refused
Breath Volume	0.000	04:04	
Air Blank	0.000	04:04	
Control Test	0.079	04:04	
Air Blank	0.000	04:05	
Diagnostics Check	OK	04:05	

Data Download Date/Time: 03/31/2016 22:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000738

INSTRUMENT REGISTERED TO: FT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOLL, MICHEL,
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: WOLL, MICHAEL
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 02/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A1
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:32

SUBJECT'S LAST NAME: HAYNES

DATE/TIME OF TEST: 03/17/2016 03:59:46

Diagnostics Check	OK	03:56	
Air Blank	0.000	03:57	
Control Test	0.080	03:57	
Air Blank	0.000	03:58	
Subject Sample	# 1	03:58	Subject Test Refused
Breath Volume	0.000	03:58	
Air Blank	0.000	03:58	
Control Test	0.078	03:59	
Air Blank	0.000	03:59	
Diagnostics Check	OK	03:59	

Data Download Date/Time: 03/31/2016 22:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000738

INSTRUMENT REGISTERED TO: FT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREEN,DONALD,H
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: GREEN,DONALD
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 02/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A1
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:18

SUBJECT'S LAST NAME: BUHLER

DATE/TIME OF TEST: 03/18/2016 05:52:28

Diagnostics Check	OK	05:44
Air Blank	0.000	05:44
Control Test	0.080	05:45
Air Blank	0.000	05:45
Subject Sample # 1	0.133	05:47
Breath Volume	2.253	05:47
Air Blank	0.000	05:48
Air Blank	0.000	05:50
Subject Sample # 2	0.132	05:50
Breath Volume	2.187	05:50
Air Blank	0.000	05:51
Control Test	0.078	05:51
Air Blank	0.000	05:52
Diagnostics Check	OK	05:52

Data Download Date/Time: 03/31/2016 22:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000738

INSTRUMENT REGISTERED TO: FT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAIN,DANIEL,A
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: CAIN,DANIEL
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 02/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A1
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:21

SUBJECT'S LAST NAME: ROBBINS

DATE/TIME OF TEST: 03/18/2016 22:51:06

Diagnostics Check	OK	22:47	
Air Blank	0.000	22:47	
Control Test	0.080	22:48	
Air Blank	0.000	22:48	
Subject Sample	# 1	22:49	Subject Test Refused
Breath Volume	0.000	22:49	
Air Blank	0.000	22:50	
Control Test	0.080	22:50	
Air Blank	0.000	22:51	
Diagnostics Check	OK	22:51	

Data Download Date/Time: 03/31/2016 22:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000738

INSTRUMENT REGISTERED TO: FT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAIN,DANIEL,A
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: CAIN,DANIEL
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 02/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A1
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:53

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 03/19/2016 05:27:39

Diagnostics Check	OK	05:19
Air Blank	0.000	05:20
Control Test	0.079	05:20
Air Blank	0.000	05:21
Subject Sample # 1	0.180	05:23
Breath Volume	3.703	05:23
Air Blank	0.000	05:23
Air Blank	0.000	05:25
Subject Sample # 2	0.172	05:26
Breath Volume	2.992	05:26
Air Blank	0.000	05:26
Control Test	0.080	05:27
Air Blank	0.000	05:27
Diagnostics Check	OK	05:27

Data Download Date/Time: 03/31/2016 22:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000738

INSTRUMENT REGISTERED TO: FT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAILOR,TRAVIS,C
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: SAILOR,TRAVIS
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 02/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A1
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:48

SUBJECT'S LAST NAME: CHANDLER

DATE/TIME OF TEST: 03/20/2016 00:43:51

Diagnostics Check	OK	00:41	
Air Blank	0.000	00:41	
Control Test	0.081	00:41	
Air Blank	0.000	00:42	
Subject Sample	# 1	00:42	Subject Test Refused
Breath Volume	0.000	00:42	
Air Blank	0.000	00:42	
Control Test	0.080	00:43	
Air Blank	0.000	00:43	
Diagnostics Check	OK	00:43	

Data Download Date/Time: 03/31/2016 22:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000738

INSTRUMENT REGISTERED TO: FT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRAWFORD,BENJAMIN,D
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: WAGNER,ROBERT
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 02/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A1
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 14:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:30

SUBJECT'S LAST NAME: DODD

DATE/TIME OF TEST: 03/22/2016 16:06:14

Diagnostics Check	OK	16:03	
Air Blank	0.000	16:03	
Control Test	0.081	16:04	
Air Blank	0.000	16:04	
Subject Sample	# 1	16:04	Subject Test Refused
Breath Volume	0.000	16:04	
Air Blank	0.000	16:05	
Control Test	0.080	16:05	
Air Blank	0.000	16:06	
Diagnostics Check	OK	16:06	

Data Download Date/Time: 03/31/2016 22:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000738

INSTRUMENT REGISTERED TO: FT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSENBURG,DUSTIN,L
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: ROSENBURG,DUSTIN
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 02/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A1
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:58

SUBJECT'S LAST NAME: ROGERS

DATE/TIME OF TEST: 03/26/2016 01:23:37

Diagnostics Check	OK	01:19	
Air Blank	0.000	01:20	
Control Test	0.080	01:20	
Air Blank	0.000	01:20	
Subject Sample	# 1	01:22	Subject Test Refused
Breath Volume	0.000	01:22	
Air Blank	0.000	01:22	
Control Test	0.080	01:23	
Air Blank	0.000	01:23	
Diagnostics Check	OK	01:23	

Data Download Date/Time: 03/31/2016 22:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUSSAK,KEVIN,M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: RUSSAK,KEVIN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 02/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:00

SUBJECT'S LAST NAME: ODOM

DATE/TIME OF TEST: 03/02/2016 05:40:04

Diagnostics Check	OK	05:32
Air Blank	0.000	05:33
Control Test	0.079	05:33
Air Blank	0.000	05:34
Subject Sample # 1	0.182	05:35
Breath Volume	3.183	05:35
Air Blank	0.000	05:35
Air Blank	0.000	05:37
Subject Sample # 2	0.175	05:38
Breath Volume	1.730	05:38
Air Blank	0.000	05:39
Control Test	0.079	05:39
Air Blank	0.000	05:39
Diagnostics Check	OK	05:40

Data Download Date/Time: 03/29/2016 10:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUSSAK,KEVIN,M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: RUSSAK,KEVIN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 02/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: HOWE

DATE/TIME OF TEST: 03/04/2016 00:38:55

Diagnostics Check	OK	00:31
Air Blank	0.000	00:31
Control Test	0.079	00:32
Air Blank	0.000	00:32
Subject Sample # 1	0.136	00:33
Breath Volume	1.894	00:33
Air Blank	0.000	00:33
Air Blank	0.000	00:35
Subject Sample # 2	0.130	00:37
Breath Volume	1.828	00:37
Air Blank	0.000	00:38
Control Test	0.080	00:38
Air Blank	0.000	00:38
Diagnostics Check	OK	00:38

Data Download Date/Time: 03/29/2016 10:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENNIS,CASEY,G
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: DENNIS,CASEY
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 02/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:12

SUBJECT'S LAST NAME: SOPHIN

DATE/TIME OF TEST: 03/09/2016 21:40:06

Diagnostics Check	OK	21:33
Air Blank	0.000	21:33
Control Test	0.078	21:33
Air Blank	0.000	21:34
Subject Sample # 1	0.137	21:35
Breath Volume	1.644	21:35
Air Blank	0.000	21:36
Air Blank	0.000	21:37
Subject Sample # 2	0.137	21:38
Breath Volume	2.582	21:38
Air Blank	0.000	21:39
Control Test	0.077	21:39
Air Blank	0.000	21:40
Diagnostics Check	OK	21:40

Data Download Date/Time: 03/29/2016 10:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FORDHAM,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: FORDHAM,CHRISTOPHER
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 02/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 15:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:40

SUBJECT'S LAST NAME: LAYTON

DATE/TIME OF TEST: 03/17/2016 17:07:24

Diagnostics Check	OK	17:01
Air Blank	0.000	17:01
Control Test	0.078	17:01
Air Blank	0.000	17:02
Subject Sample # 1	0.198	17:02
Breath Volume	4.242	17:02
Air Blank	0.000	17:03
Air Blank	0.000	17:05
Subject Sample # 2	0.180	17:05
Breath Volume	2.988	17:05
Air Blank	0.000	17:06
Control Test	0.077	17:06
Air Blank	0.000	17:07
Diagnostics Check	OK	17:07

Data Download Date/Time: 03/29/2016 10:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANE,AMANDA,M
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: LANE,AMANDA
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 02/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:49

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 03/26/2016 01:17:14

Diagnostics Check	OK	01:09	
Air Blank	0.000	01:10	
Control Test	0.078	01:10	
Air Blank	0.000	01:10	
Subject Sample # 1	0.292	01:12	
Breath Volume	1.691	01:12	
Air Blank	0.000	01:12	
Air Blank	0.000	01:14	
Subject Sample # 2	0.287	01:15	
Breath Volume	2.742	01:15	
Air Blank	0.000	01:16	
Control Test	0.074	01:16	Control Outside Tolerance
Air Blank	0.000	01:17	

Data Download Date/Time: 03/29/2016 10:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANE,AMANDA,M
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: LANE,AMANDA
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 02/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:49

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 03/26/2016 01:27:17

Diagnostics Check	OK	01:20
Air Blank	0.000	01:21
Control Test	0.078	01:21
Air Blank	0.000	01:22
Subject Sample # 1	0.294	01:22
Breath Volume	2.328	01:22
Air Blank	0.000	01:23
Air Blank	0.000	01:25
Subject Sample # 2	0.285	01:25
Breath Volume	2.613	01:25
Air Blank	0.000	01:26
Control Test	0.077	01:26
Air Blank	0.000	01:27
Diagnostics Check	OK	01:27

Data Download Date/Time: 03/29/2016 10:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANE,AMANDA,M
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: KALTZ,CAMERON
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 02/29/2016

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:17

SUBJECT'S LAST NAME: PATTON

DATE/TIME OF TEST: 03/26/2016 05:44:37

Diagnostics Check	OK	05:37
Air Blank	0.000	05:38
Control Test	0.078	05:38
Air Blank	0.000	05:39
Subject Sample # 1	0.145	05:40
Breath Volume	2.765	05:40
Air Blank	0.000	05:40
Air Blank	0.000	05:42
Subject Sample # 2	0.155	05:43
Breath Volume	2.371	05:43
Air Blank	0.000	05:43
Control Test	0.077	05:44
Air Blank	0.000	05:44
Diagnostics Check	OK	05:44

Data Download Date/Time: 03/29/2016 10:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,GLENN,H
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: ORDOQUI,DANIEL
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 02/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 06:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:27

SUBJECT'S LAST NAME: HIGHSMITH

DATE/TIME OF TEST: 03/27/2016 07:56:47

Diagnostics Check	OK	07:47	
Air Blank	0.000	07:48	
Control Test	0.078	07:48	
Air Blank	0.000	07:49	
Subject Sample	# 1	07:52	No Sample Provided
Breath Volume	0.000	07:52	
Air Blank	0.000	07:52	
Air Blank	0.000	07:54	
Subject Sample	# 2	07:55	Subject Test Refused
Breath Volume	0.000	07:55	
Air Blank	0.000	07:55	
Control Test	0.078	07:56	
Air Blank	0.000	07:56	
Diagnostics Check	OK	07:56	

Data Download Date/Time: 03/29/2016 10:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEEKS, JEREMIAH, S
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: MEEKS, JEREMIAH
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 03/27/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:42

SUBJECT'S LAST NAME: EDWARDS

DATE/TIME OF TEST: 03/29/2016 02:09:40

Diagnostics Check	OK	02:02
Air Blank	0.000	02:03
Control Test	0.079	02:03
Air Blank	0.000	02:04
Subject Sample # 1	0.194	02:04
Breath Volume	1.433	02:04
Air Blank	0.000	02:05
Air Blank	0.000	02:07
Subject Sample # 2	0.191	02:08
Breath Volume	2.128	02:08
Air Blank	0.000	02:08
Control Test	0.078	02:09
Air Blank	0.000	02:09
Diagnostics Check	OK	02:09

Data Download Date/Time: 04/13/2016 13:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK,AMANDA,R
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BUSTO,CHRISTOPHER
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 03/27/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 10:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:49

SUBJECT'S LAST NAME: KIRK

DATE/TIME OF TEST: 03/29/2016 11:47:07

Diagnostics Check	OK	11:40
Air Blank	0.000	11:40
Control Test	0.078	11:40
Air Blank	0.000	11:41
Subject Sample # 1	0.000	11:41
Breath Volume	1.574	11:41
Air Blank	0.000	11:42
Air Blank	0.000	11:44
Subject Sample # 2	0.000	11:45
Breath Volume	1.554	11:45
Air Blank	0.000	11:46
Control Test	0.079	11:46
Air Blank	0.000	11:47
Diagnostics Check	OK	11:47

Data Download Date/Time: 04/13/2016 13:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,BARBI,J
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: MYERS,BARBI
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 03/27/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 06:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:00

SUBJECT'S LAST NAME: PALMER

DATE/TIME OF TEST: 03/30/2016 10:26:39

Diagnostics Check	OK	10:20
Air Blank	0.000	10:21
Control Test	0.078	10:21
Air Blank	0.000	10:21
Subject Sample # 1	0.000	10:22
Breath Volume	2.406	10:22
Air Blank	0.000	10:22
Air Blank	0.000	10:24
Subject Sample # 2	0.000	10:25
Breath Volume	1.238	10:25
Air Blank	0.000	10:25
Control Test	0.078	10:26
Air Blank	0.000	10:26
Diagnostics Check	OK	10:26

Data Download Date/Time: 04/13/2016 13:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSE,JOHN,F
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: TAMPARY,ANTHONY
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 03/27/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 11:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:21

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 03/31/2016 12:53:49

Diagnostics Check	OK	12:47
Air Blank	0.000	12:47
Control Test	0.079	12:48
Air Blank	0.000	12:48
Subject Sample # 1	0.000	12:49
Breath Volume	4.484	12:49
Air Blank	0.000	12:49
Air Blank	0.000	12:51
Subject Sample # 2	0.000	12:52
Breath Volume	5.265	12:52
Air Blank	0.000	12:52
Control Test	0.079	12:53
Air Blank	0.000	12:53
Diagnostics Check	OK	12:53

Data Download Date/Time: 04/13/2016 13:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEVENS,MITCHELL,M
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: CHAMLIS,STEFAN
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 03/27/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:13

SUBJECT'S LAST NAME: SARASUA

DATE/TIME OF TEST: 03/31/2016 23:37:57

Diagnostics Check	OK	23:35	
Air Blank	0.000	23:35	
Control Test	0.077	23:35	
Air Blank	0.000	23:36	
Subject Sample	# 1	23:36	Subject Test Refused
Breath Volume	0.000	23:36	
Air Blank	0.000	23:37	
Control Test	0.077	23:37	
Air Blank	0.000	23:37	
Diagnostics Check	OK	23:37	

Data Download Date/Time: 04/13/2016 13:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000752

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AUSTIN,NICHOLAS,J
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: AUSTIN,NICHOLAS
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 02/17/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1766638
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: GILMET

DATE/TIME OF TEST: 03/02/2016 01:38:07

Diagnostics Check	OK	01:32
Air Blank	0.000	01:32
Control Test	0.081	01:32
Air Blank	0.000	01:33
Subject Sample # 1	0.160	01:33
Breath Volume	1.992	01:33
Air Blank	0.000	01:34
Air Blank	0.000	01:36
Subject Sample # 2	0.164	01:36
Breath Volume	2.742	01:36
Air Blank	0.000	01:37
Control Test	0.078	01:37
Air Blank	0.000	01:38
Diagnostics Check	OK	01:38

Data Download Date/Time: 03/24/2016 23:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000752

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SERRANO,ELIJHA,
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: MASLAN,JESSE
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 02/17/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1766638
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:28

SUBJECT'S LAST NAME: DAUGHERTY

DATE/TIME OF TEST: 03/02/2016 05:02:26

Diagnostics Check	OK	04:49	
Air Blank	0.000	04:49	
Control Test	0.081	04:49	
Air Blank	0.000	04:50	
Subject Sample	# 1 0.111	04:51	
Breath Volume	1.289	04:51	
Air Blank	0.000	04:51	
Air Blank	0.000	04:53	
Subject Sample	# 2 0.101	04:56	Volume Not Met
Breath Volume	0.433	04:56	
Air Blank	0.000	04:57	
Air Blank	0.000	04:59	
Subject Sample	# 3 0.110	05:00	
Breath Volume	1.140	05:00	
Air Blank	0.000	05:01	
Control Test	0.076	05:01	
Air Blank	0.000	05:02	
Diagnostics Check	OK	05:02	

Data Download Date/Time: 03/24/2016 23:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000752

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,GARY,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: MILLER,GARY
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 02/17/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1766638
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: EDDY

DATE/TIME OF TEST: 03/05/2016 01:07:26

Diagnostics Check	OK	01:00
Air Blank	0.000	01:01
Control Test	0.080	01:01
Air Blank	0.000	01:02
Subject Sample # 1	0.122	01:02
Breath Volume	1.460	01:02
Air Blank	0.000	01:03
Air Blank	0.000	01:05
Subject Sample # 2	0.116	01:05
Breath Volume	1.769	01:05
Air Blank	0.000	01:06
Control Test	0.080	01:06
Air Blank	0.000	01:07
Diagnostics Check	OK	01:07

Data Download Date/Time: 03/24/2016 23:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000752

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GAGO-
CHUQUILLANQUI,ALBERTO,E
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: GAGO-
CHUQUILLANQUI,ALBERTO
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 02/17/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1766638
DRY GAS EXPIRY DATE: 11/05/2017

VIOLATION CODE: DUI
ARREST TIME: 01:24
GENDER: Male

OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: KENNEDY

DATE/TIME OF TEST: 03/05/2016 02:52:09

Diagnostics Check	OK	02:46
Air Blank	0.000	02:46
Control Test	0.080	02:46
Air Blank	0.000	02:47
Subject Sample # 1	0.000	02:47
Breath Volume	1.500	02:47
Air Blank	0.000	02:48
Air Blank	0.000	02:50
Subject Sample # 2	0.000	02:50
Breath Volume	1.585	02:50
Air Blank	0.000	02:51
Control Test	0.079	02:51
Air Blank	0.000	02:52
Diagnostics Check	OK	02:52

Data Download Date/Time: 03/24/2016 23:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000752

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MANGUM,RACHEL,E
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: MANGUM,RACHEL
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 02/17/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1766638
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:12

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 03/05/2016 03:40:17

Diagnostics Check	OK	03:32
Air Blank	0.000	03:33
Control Test	0.080	03:33
Air Blank	0.000	03:34
Subject Sample # 1	0.111	03:35
Breath Volume	1.582	03:35
Air Blank	0.000	03:36
Air Blank	0.000	03:38
Subject Sample # 2	0.117	03:38
Breath Volume	2.515	03:38
Air Blank	0.000	03:39
Control Test	0.077	03:39
Air Blank	0.000	03:40
Diagnostics Check	OK	03:40

Data Download Date/Time: 03/24/2016 23:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000752

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOOREHEAD,DAVID,L
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: MOOREHEAD,DAVID
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 02/17/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1766638
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: ELWOOD

DATE/TIME OF TEST: 03/06/2016 04:12:16

Diagnostics Check	OK	04:05
Air Blank	0.000	04:05
Control Test	0.079	04:06
Air Blank	0.000	04:06
Subject Sample # 1	0.213	04:07
Breath Volume	1.750	04:07
Air Blank	0.000	04:08
Air Blank	0.000	04:09
Subject Sample # 2	0.207	04:10
Breath Volume	1.632	04:10
Air Blank	0.000	04:11
Control Test	0.080	04:11
Air Blank	0.000	04:12
Diagnostics Check	OK	04:12

Data Download Date/Time: 03/24/2016 23:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000752

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOBOSIEWICZ,ANTHONY,N
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: DOBOSIEWICZ,ANTHONY
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 02/17/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1766638
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:02

SUBJECT'S LAST NAME: MORENO

DATE/TIME OF TEST: 03/07/2016 23:09:55

Diagnostics Check		OK	23:01
Air Blank		0.000	23:02
Control Test		0.080	23:02
Air Blank		0.000	23:03
Subject Sample	# 1	0.236	23:04
Breath Volume		1.210	23:04
Air Blank		0.000	23:05
Air Blank		0.000	23:07
Subject Sample	# 2	0.242	23:08
Breath Volume		1.683	23:08
Air Blank		0.000	23:08
Control Test		0.078	23:09
Air Blank		0.000	23:09
Diagnostics Check		OK	23:09

Data Download Date/Time: 03/24/2016 23:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000752

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HILLIARD,JEFFERY,A
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: HILLIARD,JEFFERY
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 02/17/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1766638
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 16:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:30

SUBJECT'S LAST NAME: HARTE

DATE/TIME OF TEST: 03/09/2016 17:58:59

Diagnostics Check	OK	17:52
Air Blank	0.000	17:52
Control Test	0.080	17:53
Air Blank	0.000	17:53
Subject Sample # 1	0.280	17:54
Breath Volume	1.882	17:54
Air Blank	0.000	17:55
Air Blank	0.000	17:56
Subject Sample # 2	0.277	17:57
Breath Volume	2.960	17:57
Air Blank	0.000	17:58
Control Test	0.077	17:58
Air Blank	0.000	17:58
Diagnostics Check	OK	17:58

Data Download Date/Time: 03/24/2016 23:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000752

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AUSTIN,NICHOLAS,J
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: AUSTIN,NICHLAS
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 02/17/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1766638
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: ANDREWS

DATE/TIME OF TEST: 03/10/2016 01:31:53

Diagnostics Check	OK	01:25
Air Blank	0.000	01:26
Control Test	0.080	01:26
Air Blank	0.000	01:26
Subject Sample # 1	0.117	01:27
Breath Volume	2.332	01:27
Air Blank	0.000	01:28
Air Blank	0.000	01:29
Subject Sample # 2	0.114	01:30
Breath Volume	1.750	01:30
Air Blank	0.000	01:30
Control Test	0.079	01:31
Air Blank	0.000	01:31
Diagnostics Check	OK	01:31

Data Download Date/Time: 03/24/2016 23:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000752

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERNANDEZ,ALAIN,
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: FERNANDEZ,ALAIN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 02/17/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1766638
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 15:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:49

SUBJECT'S LAST NAME: HUDDY

DATE/TIME OF TEST: 03/10/2016 16:16:39

Diagnostics Check	OK	16:09
Air Blank	0.000	16:10
Control Test	0.079	16:10
Air Blank	0.000	16:11
Subject Sample # 1	0.067	16:12
Breath Volume	4.285	16:12
Air Blank	0.000	16:12
Air Blank	0.000	16:14
Subject Sample # 2	0.070	16:15
Breath Volume	3.406	16:15
Air Blank	0.000	16:15
Control Test	0.079	16:16
Air Blank	0.000	16:16
Diagnostics Check	OK	16:16

Data Download Date/Time: 03/24/2016 23:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000752

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALLER,JESSICA,G
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: GALLER,JESSICA
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 02/17/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1766638
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: RUSSELL

DATE/TIME OF TEST: 03/12/2016 00:37:25

Diagnostics Check	OK	00:25	
Air Blank	0.000	00:25	
Control Test	0.081	00:26	
Air Blank	0.000	00:26	
Subject Sample	# 1	00:29	No Sample Provided
Breath Volume	0.000	00:29	
Air Blank	0.000	00:30	
Air Blank	0.000	00:32	
Subject Sample	# 2	0.201	00:32
Breath Volume	1.468	00:32	
Air Blank	0.000	00:33	
Air Blank	0.000	00:35	
Subject Sample	# 3	0.195	00:35
Breath Volume	1.492	00:35	
Air Blank	0.000	00:36	
Control Test	0.079	00:36	
Air Blank	0.000	00:37	
Diagnostics Check	OK	00:37	

Data Download Date/Time: 03/24/2016 23:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000752

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAGNER,DANIEL,J
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: WAGNER,DANIEL
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 02/17/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1766638
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 17:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:39

SUBJECT'S LAST NAME: LUNGARI

DATE/TIME OF TEST: 03/17/2016 18:26:27

Diagnostics Check	OK	18:20
Air Blank	0.000	18:20
Control Test	0.080	18:20
Air Blank	0.000	18:21
Subject Sample # 1	0.018	18:22
Breath Volume	2.269	18:22
Air Blank	0.000	18:22
Air Blank	0.000	18:24
Subject Sample # 2	0.018	18:24
Breath Volume	2.164	18:24
Air Blank	0.000	18:25
Control Test	0.079	18:25
Air Blank	0.000	18:26
Diagnostics Check	OK	18:26

Data Download Date/Time: 03/24/2016 23:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000752

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LETSON,REBECCA,G
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: LETSON,REBECCA
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 02/17/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1766638
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: CASTALDI

DATE/TIME OF TEST: 03/18/2016 02:46:52

Diagnostics Check	OK	02:36	
Air Blank	0.000	02:36	
Control Test	0.081	02:36	
Air Blank	0.000	02:37	
Subject Sample	# 1	02:40	No Sample Provided
Breath Volume	0.000	02:40	
Air Blank	0.000	02:41	
Air Blank	0.000	02:43	
Subject Sample	# 2	02:45	Subject Test Refused
Breath Volume	0.000	02:45	
Air Blank	0.000	02:45	
Control Test	0.080	02:46	
Air Blank	0.000	02:46	
Diagnostics Check	OK	02:46	

Data Download Date/Time: 03/24/2016 23:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000752

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LETSON,REBECCA,G
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: LETSON,REBECCA
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 02/17/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1766638
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: CASTALDI

DATE/TIME OF TEST: 03/18/2016 03:01:44

Diagnostics Check	OK	02:52
Air Blank	0.000	02:52
Control Test	0.081	02:53
Air Blank	0.000	02:53
Subject Sample # 1	0.131	02:56
Breath Volume	2.074	02:56
Air Blank	0.000	02:57
Air Blank	0.000	02:58
Subject Sample # 2	0.120	03:00
Breath Volume	1.824	03:00
Air Blank	0.000	03:00
Control Test	0.078	03:01
Air Blank	0.000	03:01
Diagnostics Check	OK	03:01

Data Download Date/Time: 03/24/2016 23:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000752

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VISMAN,CLAYTON,A
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: VISMAN,CLAYTON
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 02/17/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1766638
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 18:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:40

SUBJECT'S LAST NAME: MORENO

DATE/TIME OF TEST: 03/18/2016 20:10:27

Diagnostics Check	OK	20:03
Air Blank	0.000	20:04
Control Test	0.080	20:04
Air Blank	0.000	20:05
Subject Sample # 1	0.333	20:05
Breath Volume	2.347	20:05
Air Blank	0.000	20:06
Air Blank	0.000	20:08
Subject Sample # 2	0.347	20:08
Breath Volume	3.523	20:08
Air Blank	0.000	20:09
Control Test	0.079	20:09
Air Blank	0.000	20:10
Diagnostics Check	OK	20:10

Data Download Date/Time: 03/24/2016 23:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000752

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TURNER,MICHAEL,R
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: MASLAN,JESSE
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 02/17/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1766638
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: TERRY

DATE/TIME OF TEST: 03/18/2016 22:15:35

Diagnostics Check	OK	22:08
Air Blank	0.000	22:09
Control Test	0.080	22:09
Air Blank	0.000	22:10
Subject Sample # 1	0.230	22:10
Breath Volume	2.929	22:10
Air Blank	0.000	22:11
Air Blank	0.000	22:13
Subject Sample # 2	0.223	22:13
Breath Volume	3.695	22:13
Air Blank	0.000	22:14
Control Test	0.078	22:14
Air Blank	0.000	22:15
Diagnostics Check	OK	22:15

Data Download Date/Time: 03/24/2016 23:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000759

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOWMAN,JOANNA,N
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BOWMAN,JOANNA
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 03/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A3
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 17:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:40

SUBJECT'S LAST NAME: JENKINS

DATE/TIME OF TEST: 03/27/2016 18:13:34

Diagnostics Check	OK	18:07
Air Blank	0.000	18:07
Control Test	0.077	18:08
Air Blank	0.000	18:08
Subject Sample # 1	0.077	18:09
Breath Volume	1.597	18:09
Air Blank	0.000	18:09
Air Blank	0.000	18:11
Subject Sample # 2	0.069	18:12
Breath Volume	1.968	18:12
Air Blank	0.000	18:12
Control Test	0.077	18:12
Air Blank	0.000	18:13
Diagnostics Check	OK	18:13

Data Download Date/Time: 04/23/2016 21:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUALMANN,DANIEL,G
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: QUALMANN,DANIEL
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 02/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:24

SUBJECT'S LAST NAME: WELCH

DATE/TIME OF TEST: 03/01/2016 04:04:02

Diagnostics Check	OK	03:52	
Air Blank	0.000	03:53	
Control Test	0.080	03:53	
Air Blank	0.000	03:53	
Subject Sample	# 1	03:57	No Sample Provided
Breath Volume	0.000	03:57	
Air Blank	0.000	03:57	
Air Blank	0.000	03:59	
Subject Sample	# 2	04:02	No Sample Provided
Breath Volume	0.000	04:02	
Air Blank	0.000	04:03	
Control Test	0.080	04:03	
Air Blank	0.000	04:03	
Diagnostics Check	OK	04:04	

Data Download Date/Time: 03/25/2016 09:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUALMANN,DANIEL,G
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: QUALMANN,DANIEL
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 02/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 05:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:24

SUBJECT'S LAST NAME: SANDERS

DATE/TIME OF TEST: 03/05/2016 06:56:09

Diagnostics Check	OK	06:50
Air Blank	0.000	06:50
Control Test	0.079	06:50
Air Blank	0.000	06:51
Subject Sample # 1	0.175	06:51
Breath Volume	2.230	06:51
Air Blank	0.000	06:52
Air Blank	0.000	06:54
Subject Sample # 2	0.161	06:54
Breath Volume	1.250	06:54
Air Blank	0.000	06:55
Control Test	0.079	06:55
Air Blank	0.000	06:56
Diagnostics Check	OK	06:56

Data Download Date/Time: 03/25/2016 09:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPBELL,ROBERT,E
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: CAMPBELL,ROBERT
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 02/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:43

SUBJECT'S LAST NAME: FLEACE

DATE/TIME OF TEST: 03/06/2016 02:58:56

Diagnostics Check	OK	02:47	
Air Blank	0.000	02:48	
Control Test	0.079	02:48	
Air Blank	0.000	02:48	
Subject Sample	# 1	02:52	No Sample Provided
Breath Volume	0.000	02:52	
Air Blank	0.000	02:52	
Air Blank	0.000	02:54	
Subject Sample	# 2	02:57	No Sample Provided
Breath Volume	0.000	02:57	
Air Blank	0.000	02:58	
Control Test	0.080	02:58	
Air Blank	0.000	02:58	
Diagnostics Check	OK	02:58	

Data Download Date/Time: 03/25/2016 09:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPBELL,ROBERT,E
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: PRESNELL,ETHAN
ARRESTING OFFICER AGENCY: High Springs PD
LAST AGENCY INSPECTION: 02/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:51

SUBJECT'S LAST NAME: YOUNGBLOOD

DATE/TIME OF TEST: 03/06/2016 03:16:32

Diagnostics Check	OK	03:15	
Air Blank	0.000	03:15	
Control Test	0.079	03:15	
Air Blank		03:16	RFI Detect
Air Blank	0.000	03:16	

Data Download Date/Time: 03/25/2016 09:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPBELL,ROBERT,E
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: PRESNELL,ETHAN
ARRESTING OFFICER AGENCY: High Springs PD
LAST AGENCY INSPECTION: 02/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:51

SUBJECT'S LAST NAME: YOUNGBLOOD

DATE/TIME OF TEST: 03/06/2016 03:26:35

Diagnostics Check	OK	03:19
Air Blank	0.000	03:19
Control Test	0.079	03:20
Air Blank	0.000	03:20
Subject Sample # 1	0.180	03:21
Breath Volume	1.257	03:21
Air Blank	0.000	03:22
Air Blank	0.000	03:24
Subject Sample # 2	0.197	03:24
Breath Volume	2.812	03:24
Air Blank	0.000	03:25
Control Test	0.077	03:25
Air Blank	0.000	03:26
Diagnostics Check	OK	03:26

Data Download Date/Time: 03/25/2016 09:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUALMANN,DANIEL,G
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: QUALMANN,DANIEL
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 02/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:16

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 03/06/2016 04:43:45

Diagnostics Check		OK	04:36
Air Blank		0.000	04:36
Control Test		0.078	04:37
Air Blank		0.000	04:37
Subject Sample	# 1	0.236	04:38
Breath Volume		2.406	04:38
Air Blank		0.000	04:39
Air Blank		0.000	04:41
Subject Sample	# 2	0.228	04:42
Breath Volume		2.835	04:42
Air Blank		0.000	04:42
Control Test		0.078	04:43
Air Blank		0.000	04:43
Diagnostics Check		OK	04:43

Data Download Date/Time: 03/25/2016 09:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUALMANN,DANIEL,G
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: NOBLES,JOHN
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 02/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:48

SUBJECT'S LAST NAME: BRADLEY

DATE/TIME OF TEST: 03/06/2016 05:15:52

Diagnostics Check	OK	05:09
Air Blank	0.000	05:09
Control Test	0.079	05:09
Air Blank	0.000	05:10
Subject Sample # 1	0.202	05:11
Breath Volume	2.023	05:11
Air Blank	0.000	05:12
Air Blank	0.000	05:13
Subject Sample # 2	0.193	05:14
Breath Volume	1.914	05:14
Air Blank	0.000	05:14
Control Test	0.077	05:15
Air Blank	0.000	05:15
Diagnostics Check	OK	05:15

Data Download Date/Time: 03/25/2016 09:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICHARDSON, MICHAEL, R
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: PINO-DIAZ, VICTOR
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 02/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 18:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:25

SUBJECT'S LAST NAME: ADVINCULA

DATE/TIME OF TEST: 03/15/2016 20:02:07

Diagnostics Check	OK	19:55
Air Blank	0.000	19:56
Control Test	0.078	19:56
Air Blank	0.000	19:57
Subject Sample # 1	0.140	19:57
Breath Volume	1.550	19:57
Air Blank	0.000	19:58
Air Blank	0.000	20:00
Subject Sample # 2	0.139	20:00
Breath Volume	2.296	20:00
Air Blank	0.000	20:01
Control Test	0.078	20:01
Air Blank	0.000	20:02
Diagnostics Check	OK	20:02

Data Download Date/Time: 03/25/2016 09:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPBELL,ROBERT,E
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: CAMPBELL,ROBERT
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 02/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:38

SUBJECT'S LAST NAME: SOARES

DATE/TIME OF TEST: 03/16/2016 02:11:38

Diagnostics Check	OK	02:02
Air Blank	0.000	02:02
Control Test	0.078	02:02
Air Blank	0.000	02:03
Subject Sample # 1	0.304	02:03
Breath Volume	3.027	02:03
Air Blank	0.000	02:04
Air Blank	0.000	02:06
Subject Sample # 2	0.279	02:06
Breath Volume	2.082	02:06
Air Blank	0.000	02:07
Air Blank	0.000	02:09
Subject Sample # 3	0.293	02:09
Breath Volume	3.070	02:09
Air Blank	0.000	02:10
Control Test	0.076	02:11
Air Blank	0.000	02:11
Diagnostics Check	OK	02:11

Data Download Date/Time: 03/25/2016 09:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUNT,CHARLES,R
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: HUNT,CHARLES
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 02/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: SHORT

DATE/TIME OF TEST: 03/18/2016 04:09:06

Diagnostics Check	OK	04:02
Air Blank	0.000	04:03
Control Test	0.079	04:03
Air Blank	0.000	04:04
Subject Sample # 1	0.091	04:04
Breath Volume	3.546	04:04
Air Blank	0.000	04:05
Air Blank	0.000	04:07
Subject Sample # 2	0.096	04:07
Breath Volume	4.039	04:07
Air Blank	0.000	04:08
Control Test	0.079	04:08
Air Blank	0.000	04:09
Diagnostics Check	OK	04:09

Data Download Date/Time: 03/25/2016 09:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUALMANN,DANIEL,G
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: QUALMANN,DANIEL
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 02/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: ALI

DATE/TIME OF TEST: 03/19/2016 03:43:35

Diagnostics Check	OK	03:37
Air Blank	0.000	03:37
Control Test	0.079	03:37
Air Blank	0.000	03:38
Subject Sample # 1	0.188	03:39
Breath Volume	2.953	03:39
Air Blank	0.000	03:39
Air Blank	0.000	03:41
Subject Sample # 2	0.185	03:41
Breath Volume	2.023	03:41
Air Blank	0.000	03:42
Control Test	0.077	03:42
Air Blank	0.000	03:43
Diagnostics Check	OK	03:43

Data Download Date/Time: 03/25/2016 09:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOSTICK,JESSE,W
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: DOSE,STEVEN
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 02/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: BRATCHER

DATE/TIME OF TEST: 03/19/2016 04:40:07

Diagnostics Check	OK	04:30	
Air Blank	0.000	04:30	
Control Test	0.078	04:31	
Air Blank	0.000	04:31	
Subject Sample	# 1 0.152	04:34	Volume Not Met
Breath Volume	0.054	04:34	
Air Blank	0.000	04:35	
Air Blank	0.000	04:37	
Subject Sample	# 2 0.141	04:38	
Breath Volume	1.878	04:38	
Air Blank		04:38	RFI Detect
Air Blank	0.000	04:39	
Control Test	0.076	04:39	
Air Blank	0.000	04:40	
Diagnostics Check	OK	04:40	

Data Download Date/Time: 03/25/2016 09:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOSTICK,JESSE,W
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: DOSE,STEVE
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 02/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: BRATCHER

DATE/TIME OF TEST: 03/19/2016 04:49:17

Diagnostics Check	OK	04:41
Air Blank	0.000	04:42
Control Test	0.078	04:42
Air Blank	0.000	04:43
Subject Sample # 1	0.145	04:44
Breath Volume	2.285	04:44
Air Blank	0.000	04:45
Air Blank	0.000	04:46
Subject Sample # 2	0.144	04:47
Breath Volume	1.183	04:47
Air Blank	0.000	04:48
Control Test	0.077	04:48
Air Blank	0.000	04:49
Diagnostics Check	OK	04:49

Data Download Date/Time: 03/25/2016 09:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRIVETTE,CHRIS,A
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: PRIVETTE,CHRIS
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 02/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 13:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:20

SUBJECT'S LAST NAME: TERESINSKI

DATE/TIME OF TEST: 03/23/2016 14:52:10

Diagnostics Check	OK	14:48	
Air Blank	0.000	14:48	
Control Test	0.078	14:48	
Air Blank	0.000	14:49	
Subject Sample	# 1	14:50	Subject Test Refused
Breath Volume	0.000	14:50	
Air Blank	0.000	14:51	
Control Test	0.079	14:51	
Air Blank	0.000	14:52	
Diagnostics Check	OK	14:52	

Data Download Date/Time: 03/25/2016 09:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOFMANN, TRAVIS, A
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: HOFMANN, TRAVIS
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 03/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 17:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:36

SUBJECT'S LAST NAME: COBB

DATE/TIME OF TEST: 03/25/2016 19:09:52

Diagnostics Check	OK	19:04	
Air Blank	0.000	19:04	
Control Test	0.078	19:04	
Air Blank	0.000	19:05	
Subject Sample # 1	0.155	19:05	
Breath Volume	2.167	19:05	
Air Blank	0.000	19:06	
Air Blank	0.000	19:08	
Subject Sample # 2	0.157	19:09	
Breath Volume	1.570	19:09	
Air Blank		19:09	RFI Detect
Air Blank	0.000	19:09	

Data Download Date/Time: 04/26/2016 09:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUNT,CHARLES,R
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: HUNT,CHARLES
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 03/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: CLEMENS

DATE/TIME OF TEST: 03/28/2016 01:12:47

Diagnostics Check	OK	01:01	
Air Blank	0.000	01:01	
Control Test	0.078	01:02	
Air Blank	0.000	01:02	
Subject Sample	# 1	01:05	No Sample Provided
Breath Volume	0.000	01:05	
Air Blank	0.000	01:06	
Air Blank	0.000	01:08	
Subject Sample	# 2	01:11	No Sample Provided
Breath Volume	0.000	01:11	
Air Blank	0.000	01:11	
Control Test	0.080	01:12	
Air Blank	0.000	01:12	
Diagnostics Check	OK	01:12	

Data Download Date/Time: 04/26/2016 09:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUALMANN,DANIEL,G
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: QUALMANN,DANIEL
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 03/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:04

SUBJECT'S LAST NAME: MARCELIN

DATE/TIME OF TEST: 03/30/2016 02:44:20

Diagnostics Check	OK	02:33	
Air Blank	0.000	02:33	
Control Test	0.079	02:33	
Air Blank	0.000	02:34	
Subject Sample	# 1	02:37	No Sample Provided
Breath Volume	0.000	02:37	
Air Blank	0.000	02:37	
Air Blank	0.000	02:39	
Subject Sample	# 2	02:42	No Sample Provided
Breath Volume	0.000	02:42	
Air Blank	0.000	02:43	
Control Test	0.080	02:43	
Air Blank	0.000	02:44	
Diagnostics Check	OK	02:44	

Data Download Date/Time: 04/26/2016 09:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICHARDSON,MICHAEL,R
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: WILLIAMS,JUSTIN
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 03/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:53

SUBJECT'S LAST NAME: HANY

DATE/TIME OF TEST: 03/30/2016 05:20:33

Diagnostics Check	OK	05:14
Air Blank	0.000	05:14
Control Test	0.079	05:14
Air Blank	0.000	05:15
Subject Sample # 1	0.171	05:16
Breath Volume	1.660	05:16
Air Blank	0.000	05:16
Air Blank	0.000	05:18
Subject Sample # 2	0.169	05:18
Breath Volume	1.792	05:18
Air Blank	0.000	05:19
Control Test	0.077	05:19
Air Blank	0.000	05:20
Diagnostics Check	OK	05:20

Data Download Date/Time: 04/26/2016 09:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000770

INSTRUMENT REGISTERED TO: HAMILTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MACLAREN,DEREK,L
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: MACLAREN,DEREK
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 02/29/2016

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 06415080A2
DRY GAS EXPIRY DATE: 04/05/2017
VIOLATION CODE: DUI
ARREST TIME: 07:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:37

SUBJECT'S LAST NAME: BURKETT

DATE/TIME OF TEST: 03/25/2016 09:03:59

Diagnostics Check	OK	08:57
Air Blank	0.000	08:57
Control Test	0.078	08:58
Air Blank	0.000	08:58
Subject Sample # 1	0.106	08:59
Breath Volume	1.667	08:59
Air Blank	0.000	08:59
Air Blank	0.000	09:01
Subject Sample # 2	0.105	09:02
Breath Volume	2.500	09:02
Air Blank	0.000	09:03
Control Test	0.077	09:03
Air Blank	0.000	09:03
Diagnostics Check	OK	09:03

Data Download Date/Time: 03/31/2016 01:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000770

INSTRUMENT REGISTERED TO: HAMILTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDONALD,SAMUEL,W
BREATH TEST OPERATOR AGENCY: Jasper PD
ARRESTING OFFICER: MCDONALD,SAMUEL
ARRESTING OFFICER AGENCY: Jasper PD
LAST AGENCY INSPECTION: 02/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06415080A2
DRY GAS EXPIRY DATE: 04/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:38

SUBJECT'S LAST NAME: MCCARTY

DATE/TIME OF TEST: 03/26/2016 00:05:51

Diagnostics Check	OK	23:59
Air Blank	0.000	23:59
Control Test	0.078	23:59
Air Blank	0.000	00:00
Subject Sample # 1	0.038	00:01
Breath Volume	5.160	00:01
Air Blank	0.000	00:01
Air Blank	0.000	00:03
Subject Sample # 2	0.035	00:04
Breath Volume	4.949	00:04
Air Blank	0.000	00:04
Control Test	0.077	00:05
Air Blank	0.000	00:05
Diagnostics Check	OK	00:05

Data Download Date/Time: 03/31/2016 01:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000771

INSTRUMENT REGISTERED TO: PERRY POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARD, RONALD, E
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: WARD, RONALD
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 02/17/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG425202
DRY GAS EXPIRY DATE: 09/09/2016
VIOLATION CODE: DUI
ARREST TIME: 16:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:53

SUBJECT'S LAST NAME: SESSIONS

DATE/TIME OF TEST: 03/08/2016 17:28:30

Diagnostics Check	OK	17:20	
Air Blank	0.000	17:21	
Control Test	0.078	17:21	
Air Blank	0.000	17:21	
Subject Sample # 1	0.276	17:22	
Breath Volume	1.707	17:22	
Air Blank	0.000	17:23	
Air Blank	0.000	17:25	
Subject Sample # 2		17:26	Subject Test Refused
Breath Volume	0.644	17:26	
Air Blank	0.000	17:27	
Control Test	0.076	17:27	
Air Blank	0.000	17:28	
Diagnostics Check	OK	17:28	

Data Download Date/Time: 03/28/2016 10:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000771

INSTRUMENT REGISTERED TO: PERRY POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRIFFIN,BRUCE,H
BREATH TEST OPERATOR AGENCY: Perry PD
ARRESTING OFFICER: LIVINGSTON,KIM
ARRESTING OFFICER AGENCY: Perry PD
LAST AGENCY INSPECTION: 02/17/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG425202
DRY GAS EXPIRY DATE: 09/09/2016
VIOLATION CODE: COURT ORD
ARREST TIME: 09:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:25

SUBJECT'S LAST NAME: CARTER

DATE/TIME OF TEST: 03/18/2016 11:04:53

Diagnostics Check	OK	10:58
Air Blank	0.000	10:59
Control Test	0.078	10:59
Air Blank	0.000	11:00
Subject Sample # 1	0.000	11:00
Breath Volume	1.257	11:00
Air Blank	0.000	11:01
Air Blank	0.000	11:03
Subject Sample # 2	0.000	11:03
Breath Volume	1.414	11:03
Air Blank	0.000	11:03
Control Test	0.076	11:04
Air Blank	0.000	11:04
Diagnostics Check	OK	11:04

Data Download Date/Time: 03/28/2016 10:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000773

INSTRUMENT REGISTERED TO: LAFAYETTE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PUTNEL,JOHN,D
BREATH TEST OPERATOR AGENCY: Lafayette County SO
ARRESTING OFFICER: DAVENPORT,LUKE
ARRESTING OFFICER AGENCY: FFWCC (Lake City)
LAST AGENCY INSPECTION: 02/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12415080A5
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 15:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:57

SUBJECT'S LAST NAME: BRAUTCHECK

DATE/TIME OF TEST: 03/21/2016 16:23:44

Diagnostics Check	OK	16:22	
Air Blank		16:23	Improper Sample
Air Blank	0.000	16:23	

Data Download Date/Time: 03/28/2016 12:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000773

INSTRUMENT REGISTERED TO: LAFAYETTE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PUTNEL,JOHN,D
BREATH TEST OPERATOR AGENCY: Lafayette County SO
ARRESTING OFFICER: DAVENPORT,LUKE
ARRESTING OFFICER AGENCY: FFWCC (Lake City)
LAST AGENCY INSPECTION: 02/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12415080A5
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 15:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:57

SUBJECT'S LAST NAME: BRAUTCHECK

DATE/TIME OF TEST: 03/21/2016 16:33:04

Diagnostics Check	OK	16:26
Air Blank	0.000	16:27
Control Test	0.078	16:27
Air Blank	0.000	16:28
Subject Sample # 1	0.201	16:28
Breath Volume	2.578	16:28
Air Blank	0.000	16:29
Air Blank	0.000	16:31
Subject Sample # 2	0.200	16:31
Breath Volume	3.277	16:31
Air Blank	0.000	16:32
Control Test	0.078	16:32
Air Blank	0.000	16:32
Diagnostics Check	OK	16:33

Data Download Date/Time: 03/28/2016 12:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIRCHARD, MARK, E
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: BISHOP, TRAMPAS
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 02/26/2016

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1686772
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 21:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:21

SUBJECT'S LAST NAME: SUTTON

DATE/TIME OF TEST: 03/14/2016 22:47:48

Diagnostics Check	OK	22:41
Air Blank	0.000	22:42
Control Test	0.082	22:42
Air Blank	0.000	22:43
Subject Sample # 1	0.112	22:43
Breath Volume	1.363	22:43
Air Blank	0.000	22:44
Air Blank	0.000	22:45
Subject Sample # 2	0.102	22:46
Breath Volume	1.539	22:46
Air Blank	0.000	22:46
Control Test	0.081	22:47
Air Blank	0.000	22:47
Diagnostics Check	OK	22:47

Data Download Date/Time: 03/24/2016 20:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOWNING,STEVEN,M
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: MACLEREN,DERRICK
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 02/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1686772
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 04:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: WORLEY

DATE/TIME OF TEST: 03/18/2016 04:52:12

Diagnostics Check	OK	04:44
Air Blank	0.000	04:44
Control Test	0.082	04:45
Air Blank	0.000	04:45
Subject Sample # 1	0.154	04:47
Breath Volume	1.746	04:47
Air Blank	0.000	04:47
Air Blank	0.000	04:49
Subject Sample # 2	0.161	04:50
Breath Volume	2.605	04:50
Air Blank	0.000	04:51
Control Test	0.082	04:51
Air Blank	0.000	04:52
Diagnostics Check	OK	04:52

Data Download Date/Time: 03/24/2016 20:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLARK,J,T
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: CLARK,J
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 02/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1686772
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 12:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:34

SUBJECT'S LAST NAME: PRICE

DATE/TIME OF TEST: 03/19/2016 13:01:25

Diagnostics Check	OK	12:55
Air Blank	0.000	12:55
Control Test	0.082	12:55
Air Blank	0.000	12:56
Subject Sample # 1	0.220	12:56
Breath Volume	2.382	12:56
Air Blank	0.000	12:57
Air Blank	0.000	12:59
Subject Sample # 2	0.230	12:59
Breath Volume	1.945	12:59
Air Blank	0.000	13:00
Control Test	0.082	13:00
Air Blank	0.000	13:01
Diagnostics Check	OK	13:01

Data Download Date/Time: 03/24/2016 20:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000780

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FALGOUT,DAVID,G
BREATH TEST OPERATOR AGENCY: Suwannee County SO
ARRESTING OFFICER: MUSGROVE,DUSTIN
ARRESTING OFFICER AGENCY: Live Oak PD
LAST AGENCY INSPECTION: 02/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06415080A1
DRY GAS EXPIRY DATE: 04/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:43

SUBJECT'S LAST NAME: MORGAN

DATE/TIME OF TEST: 03/30/2016 22:19:56

Diagnostics Check	OK	22:13
Air Blank	0.000	22:14
Control Test	0.080	22:14
Air Blank	0.000	22:15
Subject Sample # 1	0.133	22:15
Breath Volume	1.964	22:15
Air Blank	0.000	22:16
Air Blank	0.000	22:17
Subject Sample # 2	0.135	22:18
Breath Volume	2.820	22:18
Air Blank	0.000	22:19
Control Test	0.080	22:19
Air Blank	0.000	22:19
Diagnostics Check	OK	22:19

Data Download Date/Time: 04/01/2016 03:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000785

INSTRUMENT REGISTERED TO: HOLMES COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PATE, THOMAS, W
BREATH TEST OPERATOR AGENCY: Holmes County SO
ARRESTING OFFICER: PATE, THOMAS
ARRESTING OFFICER AGENCY: Holmes County SO
LAST AGENCY INSPECTION: 03/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20314080A4
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 22:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: PHILLIPS

DATE/TIME OF TEST: 03/05/2016 23:50:39

Diagnostics Check	OK	23:46	
Air Blank	0.000	23:46	
Control Test	0.079	23:46	
Air Blank	0.000	23:47	
Subject Sample	# 1	23:49	Subject Test Refused
Breath Volume	0.937	23:49	
Air Blank	0.000	23:49	
Control Test	0.082	23:50	
Air Blank	0.000	23:50	
Diagnostics Check	OK	23:50	

Data Download Date/Time: 07/03/2016 16:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000785

INSTRUMENT REGISTERED TO: HOLMES COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRENCH,JASON,K
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: FRENCH,JASON
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 03/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20314080A4
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 19:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:43

SUBJECT'S LAST NAME: TANKERSLEY

DATE/TIME OF TEST: 03/25/2016 20:15:42

Diagnostics Check	OK	20:09
Air Blank	0.000	20:09
Control Test	0.079	20:09
Air Blank	0.000	20:10
Subject Sample # 1	0.108	20:10
Breath Volume	1.105	20:10
Air Blank	0.000	20:11
Air Blank	0.000	20:13
Subject Sample # 2	0.109	20:14
Breath Volume	1.312	20:14
Air Blank	0.000	20:14
Control Test	0.079	20:15
Air Blank	0.000	20:15
Diagnostics Check	OK	20:15

Data Download Date/Time: 07/03/2016 16:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCAPIN,BRANDON,L
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: KIDD,JERMEL
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 02/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: CANNON

DATE/TIME OF TEST: 03/02/2016 02:21:47

Diagnostics Check	OK	02:19	
Air Blank	0.000	02:19	
Control Test	0.078	02:19	
Air Blank	0.000	02:20	
Subject Sample	# 1	02:20	Subject Test Refused
Breath Volume	0.000	02:20	
Air Blank	0.000	02:20	
Control Test	0.078	02:21	
Air Blank	0.000	02:21	
Diagnostics Check	OK	02:21	

Data Download Date/Time: 03/15/2016 10:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCAPIN,BRANDON,L
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: SCAPIN,BRANDON
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 02/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:43

SUBJECT'S LAST NAME: ALVAREZ

DATE/TIME OF TEST: 03/05/2016 02:08:38

Diagnostics Check	OK	02:04	
Air Blank	0.000	02:04	
Control Test	0.078	02:04	
Air Blank	0.000	02:05	
Subject Sample	# 1	02:07	Subject Test Refused
Breath Volume	0.000	02:07	
Air Blank	0.000	02:07	
Control Test	0.078	02:08	
Air Blank	0.000	02:08	
Diagnostics Check	OK	02:08	

Data Download Date/Time: 03/15/2016 10:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCAPIN,BRANDON,L
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: SCAPIN,BRANDON
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 02/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:23

SUBJECT'S LAST NAME: WELLS

DATE/TIME OF TEST: 03/06/2016 01:46:05

Diagnostics Check	OK	01:43	
Air Blank	0.000	01:43	
Control Test	0.079	01:44	
Air Blank	0.000	01:44	
Subject Sample	# 1	01:44	Subject Test Refused
Breath Volume	0.000	01:44	
Air Blank	0.000	01:45	
Control Test	0.080	01:45	
Air Blank	0.000	01:46	
Diagnostics Check	OK	01:46	

Data Download Date/Time: 03/15/2016 10:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCAPIN,BRANDON,L
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: SCAPIN,BRANDON
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 02/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: CUMMINS

DATE/TIME OF TEST: 03/07/2016 02:38:49

Diagnostics Check	OK	02:36	
Air Blank	0.000	02:36	
Control Test	0.079	02:36	
Air Blank	0.000	02:37	
Subject Sample	# 1	02:37	Subject Test Refused
Breath Volume	0.000	02:37	
Air Blank	0.000	02:37	
Control Test	0.080	02:38	
Air Blank	0.000	02:38	
Diagnostics Check	OK	02:38	

Data Download Date/Time: 03/15/2016 10:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCAPIN,BRANDON,L
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: SCAPIN,BRANDON
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 03/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:56

SUBJECT'S LAST NAME: MIMS

DATE/TIME OF TEST: 03/24/2016 02:24:21

Diagnostics Check	OK	02:16	
Air Blank	0.000	02:17	
Control Test	0.079	02:17	
Air Blank	0.000	02:18	
Subject Sample	# 1 0.218	02:19	
Breath Volume	2.957	02:19	
Air Blank	0.000	02:20	
Air Blank	0.000	02:21	
Subject Sample	# 2 0.232	02:22	
Breath Volume	3.218	02:22	
Air Blank	0.000	02:23	
Control Test	0.079	02:23	
Air Blank		02:23	RFI Detect
Air Blank	0.000	02:24	

Data Download Date/Time: 04/15/2016 10:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCAPIN,BRANDON,L
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: SCAPIN,BRANDON
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 03/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:56

SUBJECT'S LAST NAME: MIMS

DATE/TIME OF TEST: 03/24/2016 02:33:13

Diagnostics Check	OK	02:26
Air Blank	0.000	02:27
Control Test	0.078	02:27
Air Blank	0.000	02:28
Subject Sample # 1	0.230	02:28
Breath Volume	2.792	02:28
Air Blank	0.000	02:29
Air Blank	0.000	02:31
Subject Sample # 2	0.217	02:31
Breath Volume	3.425	02:31
Air Blank	0.000	02:32
Control Test	0.078	02:32
Air Blank	0.000	02:33
Diagnostics Check	OK	02:33

Data Download Date/Time: 04/15/2016 10:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERSON,THOMAS,C
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: ROBERSON,THOMAS
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 03/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 16:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:57

SUBJECT'S LAST NAME: MICARI

DATE/TIME OF TEST: 03/28/2016 17:25:41

Diagnostics Check	OK	17:19
Air Blank	0.000	17:19
Control Test	0.076	17:20
Air Blank	0.000	17:20
Subject Sample # 1	0.209	17:21
Breath Volume	2.238	17:21
Air Blank	0.000	17:21
Air Blank	0.000	17:23
Subject Sample # 2	0.213	17:24
Breath Volume	3.199	17:24
Air Blank	0.000	17:24
Control Test	0.078	17:25
Air Blank	0.000	17:25
Diagnostics Check	OK	17:25

Data Download Date/Time: 04/15/2016 10:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCAPIN,BRANDON,L
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: SCAPIN,BRANDON
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 03/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: MARTIN

DATE/TIME OF TEST: 03/31/2016 02:39:06

Diagnostics Check	OK	02:30
Air Blank	0.000	02:30
Control Test	0.079	02:31
Air Blank	0.000	02:31
Subject Sample # 1	0.193	02:32
Breath Volume	2.324	02:32
Air Blank	0.000	02:33
Air Blank	0.000	02:35
Subject Sample # 2	0.180	02:37
Breath Volume	3.136	02:37
Air Blank	0.000	02:38
Control Test	0.080	02:38
Air Blank	0.000	02:39
Diagnostics Check	OK	02:39

Data Download Date/Time: 04/15/2016 10:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNWOW,DAVID,J
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: TILLET,JON
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 02/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26614080A7
DRY GAS EXPIRY DATE: 10/05/2016
VIOLATION CODE: DUI
ARREST TIME: 01:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:44

SUBJECT'S LAST NAME: SHILLING

DATE/TIME OF TEST: 03/02/2016 03:26:43

Diagnostics Check	OK	03:20
Air Blank	0.000	03:20
Control Test	0.080	03:21
Air Blank	0.000	03:21
Subject Sample # 1	0.101	03:22
Breath Volume	2.503	03:22
Air Blank	0.000	03:22
Air Blank	0.000	03:24
Subject Sample # 2	0.098	03:25
Breath Volume	2.636	03:25
Air Blank	0.000	03:25
Control Test	0.080	03:26
Air Blank	0.000	03:26
Diagnostics Check	OK	03:26

Data Download Date/Time: 03/16/2016 14:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRKUS,JOSHUA,B
BREATH TEST OPERATOR AGENCY: Fort Pierce PD
ARRESTING OFFICER: SMITH,CLERMISE
ARRESTING OFFICER AGENCY: Fort Pierce PD
LAST AGENCY INSPECTION: 02/25/2016

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 26614080A7
DRY GAS EXPIRY DATE: 10/05/2016
VIOLATION CODE: DUI
ARREST TIME: 00:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: HALL

DATE/TIME OF TEST: 03/05/2016 01:19:04

Diagnostics Check	OK	01:16	
Air Blank	0.000	01:16	
Control Test	0.080	01:16	
Air Blank	0.000	01:17	
Subject Sample	# 1	01:17	Subject Test Refused
Breath Volume	0.000	01:17	
Air Blank	0.000	01:18	
Control Test	0.079	01:18	
Air Blank	0.000	01:18	
Diagnostics Check	OK	01:19	

Data Download Date/Time: 03/16/2016 14:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PATEL,MITHIL,V
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: PATEL,MITHIL
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 02/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26614080A7
DRY GAS EXPIRY DATE: 10/05/2016
VIOLATION CODE: DUI
ARREST TIME: 22:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: RINKER JR

DATE/TIME OF TEST: 03/06/2016 23:52:28

Diagnostics Check	OK	23:46
Air Blank	0.000	23:46
Control Test	0.080	23:46
Air Blank	0.000	23:47
Subject Sample # 1	0.142	23:47
Breath Volume	3.066	23:47
Air Blank	0.000	23:48
Air Blank	0.000	23:50
Subject Sample # 2	0.138	23:50
Breath Volume	4.371	23:50
Air Blank	0.000	23:51
Control Test	0.081	23:51
Air Blank	0.000	23:52
Diagnostics Check	OK	23:52

Data Download Date/Time: 03/16/2016 14:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WRIGHT,TANYA,L
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: DAVIS,KAYCEE
ARRESTING OFFICER AGENCY: Fort Pierce PD
LAST AGENCY INSPECTION: 02/25/2016

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 26614080A7
DRY GAS EXPIRY DATE: 10/05/2016
VIOLATION CODE: DUI
ARREST TIME: 02:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:56

SUBJECT'S LAST NAME: ROBIANI

DATE/TIME OF TEST: 03/07/2016 03:30:20

Diagnostics Check		OK	03:18	
Air Blank		0.000	03:19	
Control Test		0.081	03:19	
Air Blank		0.000	03:20	
Subject Sample	# 1	0.160	03:23	Volume Not Met
Breath Volume		0.164	03:23	
Air Blank		0.000	03:23	
Air Blank		0.000	03:25	
Subject Sample	# 2	0.162	03:28	Volume Not Met
Breath Volume		0.031	03:28	
Air Blank		0.000	03:29	
Control Test		0.079	03:29	
Air Blank		0.000	03:30	
Diagnostics Check		OK	03:30	

Data Download Date/Time: 03/16/2016 14:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNOW,DAVID,J
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: SNOW,DAVID
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 02/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26614080A7
DRY GAS EXPIRY DATE: 10/05/2016
VIOLATION CODE: DUI
ARREST TIME: 21:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:43

SUBJECT'S LAST NAME: ECHAZABAL

DATE/TIME OF TEST: 03/09/2016 23:23:37

Diagnostics Check	OK	23:17
Air Blank	0.000	23:17
Control Test	0.081	23:18
Air Blank	0.000	23:18
Subject Sample # 1	0.069	23:19
Breath Volume	1.863	23:19
Air Blank	0.000	23:19
Air Blank	0.000	23:21
Subject Sample # 2	0.065	23:22
Breath Volume	1.773	23:22
Air Blank	0.000	23:22
Control Test	0.080	23:23
Air Blank	0.000	23:23
Diagnostics Check	OK	23:23

Data Download Date/Time: 03/16/2016 14:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FELIX,GEORGE,D
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: FELIX,GEORGE
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 02/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26614080A7
DRY GAS EXPIRY DATE: 10/05/2016
VIOLATION CODE: DUI
ARREST TIME: 23:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:47

SUBJECT'S LAST NAME: NAVARIJO

DATE/TIME OF TEST: 03/12/2016 00:13:56

Diagnostics Check	OK	00:07
Air Blank	0.000	00:07
Control Test	0.081	00:08
Air Blank	0.000	00:08
Subject Sample # 1	0.171	00:09
Breath Volume	2.644	00:09
Air Blank	0.000	00:10
Air Blank	0.000	00:11
Subject Sample # 2	0.162	00:12
Breath Volume	2.687	00:12
Air Blank	0.000	00:13
Control Test	0.079	00:13
Air Blank	0.000	00:13
Diagnostics Check	OK	00:13

Data Download Date/Time: 03/16/2016 14:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALVAREZ,JUAN,R
BREATH TEST OPERATOR AGENCY: Fort Pierce PD
ARRESTING OFFICER: WACHENDORFER,JEFF
ARRESTING OFFICER AGENCY: Fort Pierce PD
LAST AGENCY INSPECTION: 03/16/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26614080A7
DRY GAS EXPIRY DATE: 10/05/2016
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: WYCHE

DATE/TIME OF TEST: 03/17/2016 02:14:41

Diagnostics Check	OK	02:08
Air Blank	0.000	02:08
Control Test	0.080	02:08
Air Blank	0.000	02:09
Subject Sample # 1	0.044	02:09
Breath Volume	1.738	02:09
Air Blank	0.000	02:10
Air Blank	0.000	02:12
Subject Sample # 2	0.044	02:13
Breath Volume	1.500	02:13
Air Blank	0.000	02:13
Control Test	0.079	02:14
Air Blank	0.000	02:14
Diagnostics Check	OK	02:14

Data Download Date/Time: 04/28/2016 13:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOGAN,TODD,A
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: HOGAN,TODD
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 03/16/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26614080A7
DRY GAS EXPIRY DATE: 10/05/2016
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:11

SUBJECT'S LAST NAME: BAEZ-ORTEGA

DATE/TIME OF TEST: 03/17/2016 02:40:28

Diagnostics Check	OK	02:33
Air Blank	0.000	02:33
Control Test	0.079	02:33
Air Blank	0.000	02:34
Subject Sample # 1	0.132	02:35
Breath Volume	1.621	02:35
Air Blank	0.000	02:36
Air Blank	0.000	02:37
Subject Sample # 2	0.141	02:38
Breath Volume	1.601	02:38
Air Blank	0.000	02:39
Control Test	0.080	02:39
Air Blank	0.000	02:40
Diagnostics Check	OK	02:40

Data Download Date/Time: 04/28/2016 13:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOGAN,TODD,A
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: HOGAN,TODD
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 03/16/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26614080A7
DRY GAS EXPIRY DATE: 10/05/2016
VIOLATION CODE: DUI
ARREST TIME: 21:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: METHOT

DATE/TIME OF TEST: 03/17/2016 21:58:29

Diagnostics Check	OK	21:52
Air Blank	0.000	21:52
Control Test	0.080	21:53
Air Blank	0.000	21:53
Subject Sample # 1	0.145	21:54
Breath Volume	1.988	21:54
Air Blank	0.000	21:54
Air Blank	0.000	21:56
Subject Sample # 2	0.140	21:56
Breath Volume	1.703	21:56
Air Blank	0.000	21:57
Control Test	0.080	21:57
Air Blank	0.000	21:58
Diagnostics Check	OK	21:58

Data Download Date/Time: 04/28/2016 13:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARNER,TODD,A
BREATH TEST OPERATOR AGENCY: Fort Pierce PD
ARRESTING OFFICER: WARNER,TODD
ARRESTING OFFICER AGENCY: Fort Pierce PD
LAST AGENCY INSPECTION: 03/16/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26614080A7
DRY GAS EXPIRY DATE: 10/05/2016
VIOLATION CODE: DUI
ARREST TIME: 00:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:03

SUBJECT'S LAST NAME: MCGANN

DATE/TIME OF TEST: 03/18/2016 01:30:05

Diagnostics Check		OK	01:23
Air Blank		0.000	01:23
Control Test		0.080	01:24
Air Blank		0.000	01:24
Subject Sample	# 1	0.255	01:25
Breath Volume		5.281	01:25
Air Blank		0.000	01:26
Air Blank		0.000	01:27
Subject Sample	# 2	0.258	01:28
Breath Volume		5.710	01:28
Air Blank		0.000	01:29
Control Test		0.080	01:29
Air Blank		0.000	01:30
Diagnostics Check		OK	01:30

Data Download Date/Time: 04/28/2016 13:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRANFORD,STEPHEN,R
BREATH TEST OPERATOR AGENCY: FHP Troop K
ARRESTING OFFICER: TIMMONS,ALLAN
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 03/16/2016

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 26614080A7
DRY GAS EXPIRY DATE: 10/05/2016
VIOLATION CODE: DUI
ARREST TIME: 22:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: CLARK

DATE/TIME OF TEST: 03/21/2016 00:07:37

Diagnostics Check		OK	00:01
Air Blank		0.000	00:01
Control Test		0.079	00:02
Air Blank		0.000	00:02
Subject Sample	# 1	0.263	00:03
Breath Volume		4.292	00:03
Air Blank		0.000	00:03
Air Blank		0.000	00:05
Subject Sample	# 2	0.262	00:06
Breath Volume		3.816	00:06
Air Blank		0.000	00:06
Control Test		0.079	00:07
Air Blank		0.000	00:07
Diagnostics Check		OK	00:07

Data Download Date/Time: 04/28/2016 13:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FELIX,GEORGE,D
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: FELIX,GEORGE
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 03/16/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26614080A7
DRY GAS EXPIRY DATE: 10/05/2016
VIOLATION CODE: DUI
ARREST TIME: 00:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: LENT

DATE/TIME OF TEST: 03/26/2016 03:15:46

Diagnostics Check	OK	03:04	
Air Blank	0.000	03:04	
Control Test	0.080	03:04	
Air Blank	0.000	03:05	
Subject Sample	# 1	03:08	No Sample Provided
Breath Volume	0.000	03:08	
Air Blank	0.000	03:09	
Air Blank	0.000	03:11	
Subject Sample	# 2 0.168	03:14	Volume Not Met
Breath Volume	0.082	03:14	
Air Blank	0.000	03:14	
Control Test	0.078	03:15	
Air Blank	0.000	03:15	
Diagnostics Check	OK	03:15	

Data Download Date/Time: 04/28/2016 13:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FELIX,GEORGE,D
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: FELIX,GEORGE
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 03/16/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26614080A7
DRY GAS EXPIRY DATE: 10/05/2016
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:42

SUBJECT'S LAST NAME: PORTER

DATE/TIME OF TEST: 03/26/2016 03:31:02

Diagnostics Check	OK	03:23
Air Blank	0.000	03:23
Control Test	0.079	03:23
Air Blank	0.000	03:24
Subject Sample # 1	0.194	03:26
Breath Volume	1.773	03:26
Air Blank	0.000	03:27
Air Blank	0.000	03:28
Subject Sample # 2	0.194	03:29
Breath Volume	1.910	03:29
Air Blank	0.000	03:30
Control Test	0.079	03:30
Air Blank	0.000	03:30
Diagnostics Check	OK	03:31

Data Download Date/Time: 04/28/2016 13:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FELIX,GEORGE,D
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: FELIX,GEORGE
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 03/16/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26614080A7
DRY GAS EXPIRY DATE: 10/05/2016
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: LENT

DATE/TIME OF TEST: 03/26/2016 03:50:18

Diagnostics Check		OK	03:35	
Air Blank		0.000	03:36	
Control Test		0.078	03:36	
Air Blank		0.000	03:37	
Subject Sample	# 1	0.173	03:40	Volume Not Met
Breath Volume		0.281	03:40	
Air Blank		0.000	03:41	
Air Blank		0.000	03:42	
Subject Sample	# 2	0.168	03:44	
Breath Volume		1.250	03:44	
Air Blank		0.000	03:44	
Air Blank		0.000	03:46	
Subject Sample	# 3	0.166	03:48	
Breath Volume		1.429	03:48	
Air Blank		0.000	03:49	
Control Test		0.079	03:49	
Air Blank		0.000	03:50	
Diagnostics Check		OK	03:50	

Data Download Date/Time: 04/28/2016 13:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARNER,TODD,A
BREATH TEST OPERATOR AGENCY: Fort Pierce PD
ARRESTING OFFICER: WARNER,TODD
ARRESTING OFFICER AGENCY: Fort Pierce PD
LAST AGENCY INSPECTION: 03/16/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26614080A7
DRY GAS EXPIRY DATE: 10/05/2016
VIOLATION CODE: DUI
ARREST TIME: 23:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: MATHIEU

DATE/TIME OF TEST: 03/26/2016 23:58:36

Diagnostics Check	OK	23:50
Air Blank	0.000	23:51
Control Test	0.079	23:51
Air Blank	0.000	23:52
Subject Sample # 1	0.217	23:54
Breath Volume	1.292	23:54
Air Blank	0.000	23:54
Air Blank	0.000	23:56
Subject Sample # 2	0.231	23:57
Breath Volume	1.355	23:57
Air Blank	0.000	23:57
Control Test	0.078	23:58
Air Blank	0.000	23:58
Diagnostics Check	OK	23:58

Data Download Date/Time: 04/28/2016 13:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FELIX,GEORGE,D
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: FELIX,GEORGE
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 03/16/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26614080A7
DRY GAS EXPIRY DATE: 10/05/2016
VIOLATION CODE: DUI
ARREST TIME: 00:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:53

SUBJECT'S LAST NAME: WARNER

DATE/TIME OF TEST: 03/27/2016 01:20:04

Diagnostics Check	OK	01:13
Air Blank	0.000	01:13
Control Test	0.080	01:14
Air Blank	0.000	01:14
Subject Sample # 1	0.190	01:15
Breath Volume	2.941	01:15
Air Blank	0.000	01:16
Air Blank	0.000	01:17
Subject Sample # 2	0.192	01:18
Breath Volume	3.136	01:18
Air Blank	0.000	01:19
Control Test	0.078	01:19
Air Blank	0.000	01:19
Diagnostics Check	OK	01:20

Data Download Date/Time: 04/28/2016 13:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALVAREZ,JUAN,R
BREATH TEST OPERATOR AGENCY: Fort Pierce PD
ARRESTING OFFICER: WACHENDORFER,JEFF
ARRESTING OFFICER AGENCY: Fort Pierce PD
LAST AGENCY INSPECTION: 03/16/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26614080A7
DRY GAS EXPIRY DATE: 10/05/2016
VIOLATION CODE: DUI
ARREST TIME: 01:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:37

SUBJECT'S LAST NAME: FOLCARELL

DATE/TIME OF TEST: 03/28/2016 02:15:57

Diagnostics Check	OK	02:09
Air Blank	0.000	02:10
Control Test	0.080	02:10
Air Blank	0.000	02:10
Subject Sample # 1	0.305	02:11
Breath Volume	2.097	02:11
Air Blank	0.000	02:12
Air Blank	0.000	02:13
Subject Sample # 2	0.324	02:14
Breath Volume	2.378	02:14
Air Blank	0.000	02:15
Control Test	0.080	02:15
Air Blank	0.000	02:15
Diagnostics Check	OK	02:15

Data Download Date/Time: 04/28/2016 13:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALVAREZ,JUAN,R
BREATH TEST OPERATOR AGENCY: Fort Pierce PD
ARRESTING OFFICER: MCCARTHY,KYLE
ARRESTING OFFICER AGENCY: Fort Pierce PD
LAST AGENCY INSPECTION: 03/16/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26614080A7
DRY GAS EXPIRY DATE: 10/05/2016
VIOLATION CODE: DUI
ARREST TIME: 02:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:07

SUBJECT'S LAST NAME: BAKER

DATE/TIME OF TEST: 03/30/2016 03:21:58

Diagnostics Check	OK	03:15
Air Blank	0.000	03:15
Control Test	0.081	03:16
Air Blank	0.000	03:16
Subject Sample # 1	0.071	03:17
Breath Volume	1.871	03:17
Air Blank	0.000	03:17
Air Blank	0.000	03:19
Subject Sample # 2	0.075	03:20
Breath Volume	2.105	03:20
Air Blank	0.000	03:21
Control Test	0.080	03:21
Air Blank	0.000	03:21
Diagnostics Check	OK	03:21

Data Download Date/Time: 04/28/2016 13:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000795

INSTRUMENT REGISTERED TO: GULF COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURCH,RUSSELL,L
BREATH TEST OPERATOR AGENCY: Port St. Joe PD
ARRESTING OFFICER: BURKETT,JESSE
ARRESTING OFFICER AGENCY: Port St. Joe PD
LAST AGENCY INSPECTION: 02/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: WATTS

DATE/TIME OF TEST: 03/02/2016 00:17:32

Diagnostics Check	OK	00:11
Air Blank	0.000	00:11
Control Test	0.078	00:11
Air Blank	0.000	00:12
Subject Sample # 1	0.108	00:13
Breath Volume	4.382	00:13
Air Blank	0.000	00:13
Air Blank	0.000	00:15
Subject Sample # 2	0.102	00:16
Breath Volume	3.445	00:16
Air Blank	0.000	00:16
Control Test	0.078	00:16
Air Blank	0.000	00:17
Diagnostics Check	OK	00:17

Data Download Date/Time: 03/31/2016 15:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000809

INSTRUMENT REGISTERED TO: DIXIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANUS,GLEN,A
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: GANNS,GLEN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 02/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 0901408A1
DRY GAS EXPIRY DATE: 05/01/2016
VIOLATION CODE: DUI
ARREST TIME: 19:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:21

SUBJECT'S LAST NAME: LEWIS

DATE/TIME OF TEST: 03/01/2016 20:46:46

Diagnostics Check	OK	20:43	
Air Blank	0.000	20:44	
Control Test	0.079	20:44	
Air Blank	0.000	20:45	
Subject Sample	# 1	20:45	Subject Test Refused
Breath Volume	0.000	20:45	
Air Blank	0.000	20:45	
Control Test	0.079	20:46	
Air Blank	0.000	20:46	
Diagnostics Check	OK	20:46	

Data Download Date/Time: 03/28/2016 11:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000809

INSTRUMENT REGISTERED TO: DIXIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FULTON,DARIN,M
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: FULTON,DARIN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 02/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 0901408A1
DRY GAS EXPIRY DATE: 05/01/2016
VIOLATION CODE: DUI
ARREST TIME: 00:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:59

SUBJECT'S LAST NAME: NORRIS

DATE/TIME OF TEST: 03/12/2016 01:31:25

Diagnostics Check	OK	01:21
Air Blank	0.000	01:22
Control Test	0.079	01:22
Air Blank	0.000	01:23
Subject Sample # 1	0.159	01:25
Breath Volume	1.156	01:25
Air Blank	0.000	01:25
Air Blank	0.000	01:27
Subject Sample # 2	0.162	01:29
Breath Volume	1.148	01:29
Air Blank	0.000	01:30
Control Test	0.079	01:30
Air Blank	0.000	01:31
Diagnostics Check	OK	01:31

Data Download Date/Time: 03/28/2016 11:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARPENTER,DONALD,M
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: CARPENTER,DONALD
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 03/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 83550
DRY GAS EXPIRY DATE: 09/26/2018
VIOLATION CODE: DUI
ARREST TIME: 23:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: NAYMON

DATE/TIME OF TEST: 03/25/2016 00:05:07

Diagnostics Check	OK	00:02	
Air Blank	0.000	00:02	
Control Test	0.078	00:02	
Air Blank	0.000	00:03	
Subject Sample	# 1	00:03	Subject Test Refused
Breath Volume	0.000	00:03	
Air Blank	0.000	00:04	
Control Test	0.079	00:04	
Air Blank	0.000	00:05	
Diagnostics Check	OK	00:05	

Data Download Date/Time: 04/28/2016 10:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRAWFORD,RICHARD,S
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: CRAWFORD,RICHARD
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 03/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1662060
DRY GAS EXPIRY DATE: 04/18/2017
VIOLATION CODE: DUI
ARREST TIME: 18:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:01

SUBJECT'S LAST NAME: GUTIERREZ

DATE/TIME OF TEST: 03/26/2016 19:34:12

Diagnostics Check	OK	19:27
Air Blank	0.000	19:28
Control Test	0.079	19:28
Air Blank	0.000	19:29
Subject Sample # 1	0.198	19:29
Breath Volume	2.429	19:29
Air Blank	0.000	19:30
Air Blank	0.000	19:32
Subject Sample # 2	0.192	19:32
Breath Volume	2.601	19:32
Air Blank	0.000	19:33
Control Test	0.080	19:33
Air Blank	0.000	19:34
Diagnostics Check	OK	19:34

Data Download Date/Time: 04/28/2016 10:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOBOSIEWICZ, ANTHONY, N
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: SCHULTZ, K
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 03/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 83550
DRY GAS EXPIRY DATE: 09/26/2018
VIOLATION CODE: DUI
ARREST TIME: 11:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:50

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 03/27/2016 12:18:13

Diagnostics Check	OK	12:10
Air Blank	0.000	12:10
Control Test	0.080	12:11
Air Blank	0.000	12:11
Subject Sample # 1	0.129	12:13
Breath Volume	1.468	12:13
Air Blank	0.000	12:14
Air Blank	0.000	12:15
Subject Sample # 2	0.129	12:16
Breath Volume	1.960	12:16
Air Blank	0.000	12:17
Control Test	0.080	12:17
Air Blank	0.000	12:18
Diagnostics Check	OK	12:18

Data Download Date/Time: 04/28/2016 10:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,GARY,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: MILLER,GARY
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 03/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 83550
DRY GAS EXPIRY DATE: 09/26/2018
VIOLATION CODE: DUI
ARREST TIME: 02:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:02

SUBJECT'S LAST NAME: EGYED

DATE/TIME OF TEST: 03/28/2016 03:29:02

Diagnostics Check		OK	03:22
Air Blank		0.000	03:23
Control Test		0.080	03:23
Air Blank		0.000	03:24
Subject Sample	# 1	0.183	03:24
Breath Volume		1.906	03:24
Air Blank		0.000	03:25
Air Blank		0.000	03:27
Subject Sample	# 2	0.181	03:27
Breath Volume		1.359	03:27
Air Blank		0.000	03:28
Control Test		0.080	03:28
Air Blank		0.000	03:28
Diagnostics Check		OK	03:29

Data Download Date/Time: 04/28/2016 10:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NASWORTH,JESSICA,A
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: NASWORTH,JESSICA
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 03/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 84550
DRY GAS EXPIRY DATE: 09/26/2018
VIOLATION CODE: DUI
ARREST TIME: 23:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:49

SUBJECT'S LAST NAME: MCDONALD

DATE/TIME OF TEST: 03/29/2016 00:17:06

Diagnostics Check	OK	00:09
Air Blank	0.000	00:10
Control Test	0.079	00:10
Air Blank	0.000	00:11
Subject Sample # 1	0.160	00:12
Breath Volume	3.277	00:12
Air Blank	0.000	00:13
Air Blank	0.000	00:14
Subject Sample # 2	0.154	00:15
Breath Volume	3.214	00:15
Air Blank	0.000	00:16
Control Test	0.080	00:16
Air Blank	0.000	00:17
Diagnostics Check	OK	00:17

Data Download Date/Time: 04/28/2016 10:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRANNEN,JUSTIN,C
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: WILLS,A
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 02/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28814080A3
DRY GAS EXPIRY DATE: 11/05/2016
VIOLATION CODE: DUI
ARREST TIME: 09:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:36

SUBJECT'S LAST NAME: HYSELL

DATE/TIME OF TEST: 03/09/2016 10:31:33

Diagnostics Check	OK	10:28	
Air Blank	0.000	10:29	
Control Test	0.079	10:29	
Air Blank	0.000	10:29	
Subject Sample	# 1	10:30	Subject Test Refused
Breath Volume	0.000	10:30	
Air Blank	0.000	10:30	
Control Test	0.078	10:30	
Air Blank	0.000	10:31	
Diagnostics Check	OK	10:31	

Data Download Date/Time: 03/28/2016 13:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLIS,TERENCE,L
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: TAN,RICHARD
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 02/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28814080A3
DRY GAS EXPIRY DATE: 11/05/2016
VIOLATION CODE: DUI
ARREST TIME: 22:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:23

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 03/12/2016 23:46:40

Diagnostics Check	OK	23:43	
Air Blank	0.000	23:44	
Control Test	0.078	23:44	
Air Blank	0.000	23:44	
Subject Sample	# 1	23:45	Subject Test Refused
Breath Volume	0.000	23:45	
Air Blank	0.000	23:45	
Control Test	0.078	23:46	
Air Blank	0.000	23:46	
Diagnostics Check	OK	23:46	

Data Download Date/Time: 03/28/2016 13:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SEGREST,DARRELL,E
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: TEST,TEST
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 02/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28814080A3
DRY GAS EXPIRY DATE: 11/05/2016
VIOLATION CODE: DUI
ARREST TIME: 13:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:30

SUBJECT'S LAST NAME: CARUTHERS

DATE/TIME OF TEST: 03/19/2016 14:05:26

Diagnostics Check	OK	14:04	
Air Blank	0.000	14:04	
Control Test	0.079	14:04	RFI Detect
Air Blank	0.000	14:05	

Data Download Date/Time: 03/28/2016 13:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SEGREST,DARRELL,E
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: TEST,T
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 02/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28814080A3
DRY GAS EXPIRY DATE: 11/05/2016
VIOLATION CODE: DUI
ARREST TIME: 13:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:30

SUBJECT'S LAST NAME: CARUTHERS

DATE/TIME OF TEST: 03/19/2016 14:12:43

Diagnostics Check	OK	14:06
Air Blank	0.000	14:07
Control Test	0.078	14:07
Air Blank	0.000	14:08
Subject Sample # 1	0.000	14:08
Breath Volume	1.511	14:08
Air Blank	0.000	14:08
Air Blank	0.000	14:10
Subject Sample # 2	0.000	14:11
Breath Volume	1.632	14:11
Air Blank	0.000	14:11
Control Test	0.078	14:12
Air Blank	0.000	14:12
Diagnostics Check	OK	14:12

Data Download Date/Time: 03/28/2016 13:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SEGREST,DARRELL,E
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: TEST,T
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 02/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28814080A3
DRY GAS EXPIRY DATE: 11/05/2016
VIOLATION CODE: DUI
ARREST TIME: 12:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 12:15

SUBJECT'S LAST NAME: CARUTHERS

DATE/TIME OF TEST: 03/20/2016 12:47:27

Diagnostics Check	OK	12:41
Air Blank	0.000	12:42
Control Test	0.078	12:42
Air Blank	0.000	12:42
Subject Sample # 1	0.000	12:43
Breath Volume	1.449	12:43
Air Blank	0.000	12:43
Air Blank	0.000	12:45
Subject Sample # 2	0.000	12:46
Breath Volume	1.421	12:46
Air Blank	0.000	12:46
Control Test	0.076	12:46
Air Blank	0.000	12:47
Diagnostics Check	OK	12:47

Data Download Date/Time: 03/28/2016 13:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GEOGHEGAN,ASHLEY,M
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: SMITH,BILLY
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 02/13/2016

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 28814080A3
DRY GAS EXPIRY DATE: 11/05/2016
VIOLATION CODE: DUI
ARREST TIME: 17:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:38

SUBJECT'S LAST NAME: GOLEMBESKI JR

DATE/TIME OF TEST: 03/20/2016 19:13:59

Diagnostics Check	OK	19:02	
Air Blank	0.000	19:03	
Control Test	0.079	19:03	
Air Blank	0.000	19:03	
Subject Sample	# 1	19:07	No Sample Provided
Breath Volume	0.000	19:07	
Air Blank	0.000	19:07	
Air Blank	0.000	19:09	
Subject Sample	# 2	19:12	No Sample Provided
Breath Volume	0.000	19:12	
Air Blank	0.000	19:13	
Control Test	0.078	19:13	
Air Blank	0.000	19:13	
Diagnostics Check	OK	19:13	

Data Download Date/Time: 03/28/2016 13:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GEOGHEGAN,ASHLEY,M
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: SMITH,BILLY
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 02/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28814080A3
DRY GAS EXPIRY DATE: 11/05/2016
VIOLATION CODE: DUI
ARREST TIME: 17:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:38

SUBJECT'S LAST NAME: GOLEMBESKI

DATE/TIME OF TEST: 03/20/2016 19:23:12

Diagnostics Check		OK	19:16
Air Blank		0.000	19:16
Control Test		0.079	19:17
Air Blank		0.000	19:17
Subject Sample	# 1	0.240	19:18
Breath Volume		4.132	19:18
Air Blank		0.000	19:19
Air Blank		0.000	19:20
Subject Sample	# 2	0.244	19:21
Breath Volume		3.453	19:21
Air Blank		0.000	19:22
Control Test		0.078	19:22
Air Blank		0.000	19:23
Diagnostics Check		OK	19:23

Data Download Date/Time: 03/28/2016 13:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GEOGHEGAN,ASHLEY,M
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: MATA,ROBERT
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 02/13/2016

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 28814080A3
DRY GAS EXPIRY DATE: 11/05/2016
VIOLATION CODE: DUI
ARREST TIME: 20:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:16

SUBJECT'S LAST NAME: CARONE

DATE/TIME OF TEST: 03/24/2016 21:44:02

Diagnostics Check	OK	21:37
Air Blank	0.000	21:37
Control Test	0.077	21:37
Air Blank	0.000	21:38
Subject Sample # 1	0.132	21:39
Breath Volume	2.328	21:39
Air Blank	0.000	21:40
Air Blank	0.000	21:41
Subject Sample # 2	0.133	21:42
Breath Volume	2.566	21:42
Air Blank	0.000	21:43
Control Test	0.076	21:43
Air Blank	0.000	21:43
Diagnostics Check	OK	21:44

Data Download Date/Time: 03/28/2016 13:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WING,JOSHUA,M
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: VICK,BRYAN
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 02/16/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 16:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:12

SUBJECT'S LAST NAME: GIDEON

DATE/TIME OF TEST: 03/01/2016 17:39:04

Diagnostics Check	OK	17:33
Air Blank	0.000	17:33
Control Test	0.077	17:33
Air Blank	0.000	17:34
Subject Sample # 1	0.026	17:34
Breath Volume	2.609	17:34
Air Blank	0.000	17:35
Air Blank	0.000	17:37
Subject Sample # 2	0.025	17:37
Breath Volume	1.292	17:37
Air Blank	0.000	17:38
Control Test	0.078	17:38
Air Blank	0.000	17:38
Diagnostics Check	OK	17:39

Data Download Date/Time: 03/18/2016 10:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAWRENCE,JEDIDAH,E
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: HORTON,T
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 02/16/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: MOTT

DATE/TIME OF TEST: 03/06/2016 03:45:35

Diagnostics Check	OK	03:42	
Air Blank	0.000	03:42	
Control Test	0.077	03:42	
Air Blank	0.000	03:43	
Subject Sample	# 1	03:44	Subject Test Refused
Breath Volume	0.000	03:44	
Air Blank	0.000	03:44	
Control Test	0.077	03:44	
Air Blank	0.000	03:45	
Diagnostics Check	OK	03:45	

Data Download Date/Time: 03/18/2016 10:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RHODE,ERIC,C
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: TAYLOR,T
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 02/16/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 17:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:56

SUBJECT'S LAST NAME: RICHARDS

DATE/TIME OF TEST: 03/06/2016 19:30:30

Diagnostics Check	OK	19:22
Air Blank	0.000	19:23
Control Test	0.077	19:23
Air Blank	0.000	19:24
Subject Sample # 1	0.312	19:25
Breath Volume	2.906	19:25
Air Blank	0.000	19:26
Air Blank	0.000	19:28
Subject Sample # 2	0.323	19:28
Breath Volume	2.902	19:28
Air Blank	0.000	19:29
Control Test	0.077	19:29
Air Blank	0.000	19:30
Diagnostics Check	OK	19:30

Data Download Date/Time: 03/18/2016 10:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOGUE,VINCE,H
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: MURPHY,J
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 02/16/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: WOODARD

DATE/TIME OF TEST: 03/06/2016 22:01:21

Diagnostics Check	OK	21:53	
Air Blank	0.000	21:53	
Control Test	0.077	21:54	
Air Blank	0.000	21:54	
Subject Sample	# 1 0.133	21:55	
Breath Volume	1.542	21:55	
Air Blank	0.000	21:56	
Air Blank	0.000	21:58	
Subject Sample	# 2	21:59	Subject Test Refused
Breath Volume	0.378	21:59	
Air Blank	0.000	22:00	
Control Test	0.077	22:00	
Air Blank	0.000	22:01	
Diagnostics Check	OK	22:01	

Data Download Date/Time: 03/18/2016 10:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RHODE,ERIC,C
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: HORTON,T
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 02/16/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:14

SUBJECT'S LAST NAME: YOST

DATE/TIME OF TEST: 03/07/2016 02:51:31

Diagnostics Check	OK	02:48	
Air Blank	0.000	02:49	
Control Test	0.077	02:49	
Air Blank	0.000	02:49	
Subject Sample	# 1	02:50	Subject Test Refused
Breath Volume	0.000	02:50	
Air Blank	0.000	02:50	
Control Test	0.077	02:50	
Air Blank	0.000	02:51	
Diagnostics Check	OK	02:51	

Data Download Date/Time: 03/18/2016 10:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCAZZIO,SEAN,J
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: HOWARD,R
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 02/16/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 18:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:50

SUBJECT'S LAST NAME: STEADMAN

DATE/TIME OF TEST: 03/07/2016 19:25:53

Diagnostics Check	OK	19:20	
Air Blank	0.000	19:20	
Control Test	0.077	19:20	
Air Blank	0.000	19:21	
Subject Sample # 1	0.110	19:21	
Breath Volume	1.445	19:21	
Air Blank	0.000	19:22	
Air Blank	0.000	19:24	
Subject Sample # 2	0.113	19:24	
Breath Volume	1.480	19:24	
Air Blank	0.000	19:24	
Control Test	0.074	19:25	Control Outside Tolerance
Air Blank	0.000	19:25	

Data Download Date/Time: 03/18/2016 10:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHOETTNER,ADRIAN,J
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: HOWARD,R
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 02/16/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: ANZELLOTTI

DATE/TIME OF TEST: 03/07/2016 21:58:16

Diagnostics Check	OK	21:51
Air Blank	0.000	21:52
Control Test	0.077	21:52
Air Blank	0.000	21:52
Subject Sample # 1	0.000	21:53
Breath Volume	3.042	21:53
Air Blank	0.000	21:54
Air Blank	0.000	21:56
Subject Sample # 2	0.000	21:56
Breath Volume	2.140	21:56
Air Blank	0.000	21:57
Control Test	0.077	21:57
Air Blank	0.000	21:58
Diagnostics Check	OK	21:58

Data Download Date/Time: 03/18/2016 10:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHOETTNER,ADRIAN,J
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: HOWARD,R
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 02/16/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:56

SUBJECT'S LAST NAME: COLEMAN

DATE/TIME OF TEST: 03/08/2016 02:20:20

Diagnostics Check	OK	02:17	
Air Blank	0.000	02:17	
Control Test	0.077	02:18	
Air Blank	0.000	02:18	
Subject Sample	# 1	02:18	Subject Test Refused
Breath Volume	0.000	02:18	
Air Blank	0.000	02:19	
Control Test	0.077	02:19	
Air Blank	0.000	02:20	
Diagnostics Check	OK	02:20	

Data Download Date/Time: 03/18/2016 10:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WING,JOSHUA,M
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: COUNTY TRANSIT,COUNTY
TRANSIT
ARRESTING OFFICER AGENCY: CCA Citrus
LAST AGENCY INSPECTION: 02/16/2016

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: ZERO TOL
ARREST TIME: 00:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:15

SUBJECT'S LAST NAME: CARPENTER

DATE/TIME OF TEST: 03/10/2016 09:44:25

Diagnostics Check	OK	09:38
Air Blank	0.000	09:38
Control Test	0.077	09:39
Air Blank	0.000	09:39
Subject Sample # 1	0.000	09:40
Breath Volume	1.886	09:40
Air Blank	0.000	09:40
Air Blank	0.000	09:42
Subject Sample # 2	0.000	09:43
Breath Volume	1.281	09:43
Air Blank	0.000	09:43
Control Test	0.077	09:43
Air Blank	0.000	09:44
Diagnostics Check	OK	09:44

Data Download Date/Time: 03/18/2016 10:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WING,JOSHUA,M
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: COUNTY TRANSIT,COUNTY
TRANSIT
ARRESTING OFFICER AGENCY: CCA Citrus
LAST AGENCY INSPECTION: 02/16/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: ZERO TOL
ARREST TIME: 00:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:33

SUBJECT'S LAST NAME: CAPRENTER

DATE/TIME OF TEST: 03/10/2016 09:59:13

Diagnostics Check	OK	09:53
Air Blank	0.000	09:53
Control Test	0.077	09:53
Air Blank	0.000	09:54
Subject Sample # 1	0.000	09:54
Breath Volume	1.363	09:54
Air Blank	0.000	09:55
Air Blank	0.000	09:57
Subject Sample # 2	0.000	09:57
Breath Volume	1.324	09:57
Air Blank	0.000	09:58
Control Test	0.077	09:58
Air Blank	0.000	09:59
Diagnostics Check	OK	09:59

Data Download Date/Time: 03/18/2016 10:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCAZZIO,SEAN,J
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: HOWARD,R
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 02/16/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:25

SUBJECT'S LAST NAME: STEPHENSON

DATE/TIME OF TEST: 03/11/2016 22:57:32

Diagnostics Check	OK	22:54	
Air Blank	0.000	22:55	
Control Test	0.078	22:55	
Air Blank	0.000	22:55	
Subject Sample	# 1	22:56	Subject Test Refused
Breath Volume	0.000	22:56	
Air Blank	0.000	22:56	
Control Test	0.077	22:56	
Air Blank	0.000	22:57	
Diagnostics Check	OK	22:57	

Data Download Date/Time: 03/18/2016 10:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RHODE,ERIC,C
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: HUNTER,J
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 03/18/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:07

SUBJECT'S LAST NAME: MCELMURRAY

DATE/TIME OF TEST: 03/19/2016 03:47:31

Diagnostics Check	OK	03:41
Air Blank	0.000	03:41
Control Test	0.077	03:41
Air Blank	0.000	03:42
Subject Sample # 1	0.171	03:42
Breath Volume	1.792	03:42
Air Blank	0.000	03:43
Air Blank	0.000	03:45
Subject Sample # 2	0.173	03:45
Breath Volume	1.664	03:45
Air Blank	0.000	03:46
Control Test	0.076	03:46
Air Blank	0.000	03:47
Diagnostics Check	OK	03:47

Data Download Date/Time: 04/28/2016 11:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RHODE,ERIC,C
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: HUNTER,J
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 03/18/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:01

SUBJECT'S LAST NAME: RINGLER

DATE/TIME OF TEST: 03/20/2016 00:28:04

Diagnostics Check	OK	00:25	
Air Blank	0.000	00:25	
Control Test	0.078	00:25	
Air Blank	0.000	00:26	
Subject Sample	# 1	00:26	Subject Test Refused
Breath Volume	0.000	00:26	
Air Blank	0.000	00:27	
Control Test	0.078	00:27	
Air Blank	0.000	00:27	
Diagnostics Check	OK	00:28	

Data Download Date/Time: 04/28/2016 11:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHOETTNER,ADRIAN,J
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: HOWARD,R
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 03/18/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:39

SUBJECT'S LAST NAME: SHELL

DATE/TIME OF TEST: 03/22/2016 00:02:02

Diagnostics Check	OK	23:59	
Air Blank	0.000	23:59	
Control Test	0.078	00:00	
Air Blank	0.000	00:00	
Subject Sample	# 1	00:00	Subject Test Refused
Breath Volume	0.000	00:00	
Air Blank	0.000	00:01	
Control Test	0.078	00:01	
Air Blank	0.000	00:01	
Diagnostics Check	OK	00:02	

Data Download Date/Time: 04/28/2016 11:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: METZ,KYLE,J
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: WHITSEL,TROOPER
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/18/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 08:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:58

SUBJECT'S LAST NAME: ORGERON

DATE/TIME OF TEST: 03/27/2016 12:25:12

Diagnostics Check	OK	12:22	
Air Blank	0.000	12:22	
Control Test	0.078	12:23	
Air Blank	0.000	12:23	
Subject Sample	# 1	12:23	Subject Test Refused
Breath Volume	0.000	12:23	
Air Blank	0.000	12:24	
Control Test	0.077	12:24	
Air Blank	0.000	12:25	
Diagnostics Check	OK	12:25	

Data Download Date/Time: 04/28/2016 11:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SWIGER,CARL,M
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: N/A,N/A
ARRESTING OFFICER AGENCY: OTHER County
LAST AGENCY INSPECTION: 03/18/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: OTHER
ARREST TIME: 08:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:15

SUBJECT'S LAST NAME: NUGENT

DATE/TIME OF TEST: 03/28/2016 08:43:31

Diagnostics Check	OK	08:36
Air Blank	0.000	08:37
Control Test	0.077	08:37
Air Blank	0.000	08:37
Subject Sample # 1	0.000	08:39
Breath Volume	2.367	08:39
Air Blank	0.000	08:39
Air Blank	0.000	08:41
Subject Sample # 2	0.000	08:42
Breath Volume	1.652	08:42
Air Blank	0.000	08:42
Control Test	0.078	08:42
Air Blank	0.000	08:43
Diagnostics Check	OK	08:43

Data Download Date/Time: 04/28/2016 11:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRENSHAW,SHANNON,L
BREATH TEST OPERATOR AGENCY: Leesburg PD
ARRESTING OFFICER: SCHEER,J
ARRESTING OFFICER AGENCY: Leesburg PD
LAST AGENCY INSPECTION: 03/19/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A1
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:55

SUBJECT'S LAST NAME: WATSON

DATE/TIME OF TEST: 03/19/2016 23:27:40

Diagnostics Check		OK	23:16	
Air Blank		0.000	23:16	
Control Test		0.079	23:17	
Air Blank		0.000	23:17	
Subject Sample	# 1	0.160	23:20	Volume Not Met
Breath Volume		0.074	23:20	
Air Blank		0.000	23:21	
Air Blank		0.000	23:23	
Subject Sample	# 2	0.000	23:26	Volume Not Met
Breath Volume		0.042	23:26	
Air Blank		0.000	23:26	
Control Test		0.078	23:27	
Air Blank		0.000	23:27	
Diagnostics Check		OK	23:27	

Data Download Date/Time: 04/30/2016 08:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,JONATHAN,B
BREATH TEST OPERATOR AGENCY: Leesburg PD
ARRESTING OFFICER: WILLIAMS,JONATHAN
ARRESTING OFFICER AGENCY: Leesburg PD
LAST AGENCY INSPECTION: 03/19/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A1
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: BARKER

DATE/TIME OF TEST: 03/31/2016 00:21:46

Diagnostics Check	OK	00:15
Air Blank	0.000	00:15
Control Test	0.079	00:15
Air Blank	0.000	00:16
Subject Sample # 1	0.115	00:17
Breath Volume	2.125	00:17
Air Blank	0.000	00:17
Air Blank	0.000	00:19
Subject Sample # 2	0.114	00:20
Breath Volume	2.097	00:20
Air Blank	0.000	00:20
Control Test	0.078	00:21
Air Blank	0.000	00:21
Diagnostics Check	OK	00:21

Data Download Date/Time: 04/30/2016 08:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRENSHAW,SHANNON,L
BREATH TEST OPERATOR AGENCY: Leesburg PD
ARRESTING OFFICER: CRENSHAW,S
ARRESTING OFFICER AGENCY: Leesburg PD
LAST AGENCY INSPECTION: 03/19/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A1
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 16:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:18

SUBJECT'S LAST NAME: ELDREDGE

DATE/TIME OF TEST: 03/31/2016 16:47:25

Diagnostics Check	OK	16:40
Air Blank	0.000	16:41
Control Test	0.079	16:41
Air Blank	0.000	16:41
Subject Sample # 1	0.223	16:42
Breath Volume	1.269	16:42
Air Blank	0.000	16:42
Air Blank	0.000	16:44
Subject Sample # 2	0.222	16:45
Breath Volume	2.070	16:45
Air Blank	0.000	16:46
Control Test	0.079	16:46
Air Blank	0.000	16:47
Diagnostics Check	OK	16:47

Data Download Date/Time: 04/30/2016 08:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY,PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRABAND,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/01/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:06

SUBJECT'S LAST NAME: RAMOS

DATE/TIME OF TEST: 03/01/2016 21:38:23

Diagnostics Check	OK	21:31
Air Blank	0.000	21:31
Control Test	0.079	21:32
Air Blank	0.000	21:32
Subject Sample # 1	0.213	21:33
Breath Volume	2.968	21:33
Air Blank	0.000	21:34
Air Blank	0.000	21:36
Subject Sample # 2	0.223	21:36
Breath Volume	3.492	21:36
Air Blank	0.000	21:37
Control Test	0.080	21:37
Air Blank	0.000	21:38
Diagnostics Check	OK	21:38

Data Download Date/Time: 03/03/2016 20:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARLAUG, G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/01/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:44

SUBJECT'S LAST NAME: NELSON

DATE/TIME OF TEST: 03/02/2016 23:12:44

Diagnostics Check	OK	23:04
Air Blank	0.000	23:05
Control Test	0.079	23:05
Air Blank	0.000	23:05
Subject Sample # 1	0.044	23:06
Breath Volume	1.574	23:06
Air Blank	0.000	23:07
Air Blank	0.000	23:09
Subject Sample # 2	0.043	23:11
Breath Volume	1.226	23:11
Air Blank	0.000	23:11
Control Test	0.080	23:12
Air Blank	0.000	23:12
Diagnostics Check	OK	23:12

Data Download Date/Time: 03/03/2016 20:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARLAUG,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/01/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: NORTON

DATE/TIME OF TEST: 03/03/2016 02:54:28

Diagnostics Check	OK	02:48
Air Blank	0.000	02:48
Control Test	0.079	02:49
Air Blank	0.000	02:49
Subject Sample # 1	0.092	02:49
Breath Volume	1.843	02:49
Air Blank	0.000	02:50
Air Blank	0.000	02:52
Subject Sample # 2	0.093	02:52
Breath Volume	2.054	02:52
Air Blank	0.000	02:53
Control Test	0.079	02:53
Air Blank	0.000	02:54
Diagnostics Check	OK	02:54

Data Download Date/Time: 03/03/2016 20:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 03/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 05/08/2017
VIOLATION CODE: DUI
ARREST TIME: 02:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:56

SUBJECT'S LAST NAME: CESARI

DATE/TIME OF TEST: 03/04/2016 03:22:40

Diagnostics Check		OK	03:16
Air Blank		0.000	03:16
Control Test		0.081	03:17
Air Blank		0.000	03:17
Subject Sample	# 1	0.200	03:18
Breath Volume		4.281	03:18
Air Blank		0.000	03:18
Air Blank		0.000	03:20
Subject Sample	# 2	0.200	03:21
Breath Volume		3.925	03:21
Air Blank		0.000	03:21
Control Test		0.081	03:22
Air Blank		0.000	03:22
Diagnostics Check		OK	03:22

Data Download Date/Time: 03/07/2016 12:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK,CASEY,S
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 03/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:23

SUBJECT'S LAST NAME: AVENT

DATE/TIME OF TEST: 03/05/2016 00:46:33

Diagnostics Check	Fail	00:46	DSP Test
Air Blank		00:46	Ambient Fail

Data Download Date/Time: 03/07/2016 12:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK,CASEY,S
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 03/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:23

SUBJECT'S LAST NAME: AVENT

DATE/TIME OF TEST: 03/05/2016 00:57:24

Diagnostics Check		OK	00:49
Air Blank		0.000	00:50
Control Test		0.081	00:50
Air Blank		0.000	00:51
Subject Sample	# 1	0.103	00:52
Breath Volume		1.144	00:52
Air Blank		0.000	00:53
Air Blank		0.000	00:55
Subject Sample	# 2	0.103	00:55
Breath Volume		1.371	00:55
Air Blank		0.000	00:56
Control Test		0.081	00:56
Air Blank		0.000	00:57
Diagnostics Check		OK	00:57

Data Download Date/Time: 03/07/2016 12:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMMONS,ANGELA,V
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: TINDALL,GREGORY
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 03/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 07:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:40

SUBJECT'S LAST NAME: FIGUEROA

DATE/TIME OF TEST: 03/05/2016 08:07:09

Diagnostics Check	OK	08:00
Air Blank	0.000	08:00
Control Test	0.081	08:01
Air Blank	0.000	08:01
Subject Sample # 1	0.193	08:02
Breath Volume	2.382	08:02
Air Blank	0.000	08:03
Air Blank	0.000	08:05
Subject Sample # 2	0.179	08:05
Breath Volume	1.554	08:05
Air Blank	0.000	08:06
Control Test	0.081	08:06
Air Blank	0.000	08:07
Diagnostics Check	OK	08:07

Data Download Date/Time: 03/07/2016 12:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SARRO,TAYLOR,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRIGHTWELL,R
ARRESTING OFFICER AGENCY: U of So FL Campus PD
LAST AGENCY INSPECTION: 03/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:58

SUBJECT'S LAST NAME: SIBLEY-PERONE

DATE/TIME OF TEST: 03/06/2016 00:21:54

Diagnostics Check	OK	00:18	
Air Blank	0.000	00:18	
Control Test	0.082	00:19	
Air Blank	0.000	00:19	
Subject Sample	# 1 0.193	00:20	
Breath Volume	2.570	00:20	
Air Blank		00:21	Purge Fail
Air Blank		00:21	Purge Fail

Data Download Date/Time: 03/07/2016 12:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SARRO,TAYLOR,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRIGHTWELL,R
ARRESTING OFFICER AGENCY: U of So FL Campus PD
LAST AGENCY INSPECTION: 03/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:58

SUBJECT'S LAST NAME: SIBLEY-PERONE

DATE/TIME OF TEST: 03/06/2016 00:23:34

Diagnostics Check	Fail	00:23	DSP Test
Air Blank	0.000	00:23	

Data Download Date/Time: 03/07/2016 12:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SARRO,TAYLOR,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRIGHTWELL,R
ARRESTING OFFICER AGENCY: U of So FL Campus PD
LAST AGENCY INSPECTION: 03/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:58

SUBJECT'S LAST NAME: SIBLEY-PERONE

DATE/TIME OF TEST: 03/06/2016 00:26:13

Diagnostics Check	Fail	00:25	DSP Test
Air Blank		00:26	Purge Fail

Data Download Date/Time: 03/07/2016 12:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FINDLEY,CARLTONJANE,P
BREATH TEST OPERATOR AGENCY: Groveland PD
ARRESTING OFFICER: FINDLEY,CARLTONJANE
ARRESTING OFFICER AGENCY: Groveland PD
LAST AGENCY INSPECTION: 02/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12014080A1
DRY GAS EXPIRY DATE: 06/01/2016
VIOLATION CODE: DUI
ARREST TIME: 23:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:43

SUBJECT'S LAST NAME: DELGADO

DATE/TIME OF TEST: 03/08/2016 01:15:31

Diagnostics Check	OK	01:04	
Air Blank	0.000	01:04	
Control Test	0.078	01:04	
Air Blank	0.000	01:05	
Subject Sample	# 1	01:08	No Sample Provided
Breath Volume	0.000	01:08	
Air Blank	0.000	01:09	
Air Blank	0.000	01:11	
Subject Sample	# 2	01:14	No Sample Provided
Breath Volume	0.000	01:14	
Air Blank	0.000	01:14	
Control Test	0.079	01:14	
Air Blank	0.000	01:15	
Diagnostics Check	OK	01:15	

Data Download Date/Time: 03/31/2016 22:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUPPINO,SETH,M
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: LUPPINO,SETH
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 02/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12014080A1
DRY GAS EXPIRY DATE: 06/01/2016
VIOLATION CODE: DUI
ARREST TIME: 06:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:10

SUBJECT'S LAST NAME: REAVES

DATE/TIME OF TEST: 03/19/2016 06:53:38

Diagnostics Check		OK	06:41	
Air Blank		0.000	06:42	
Control Test		0.078	06:42	
Air Blank		0.000	06:43	
Subject Sample	# 1	0.156	06:46	Volume Not Met
Breath Volume		0.097	06:46	
Air Blank		0.000	06:47	
Air Blank		0.000	06:48	
Subject Sample	# 2	0.118	06:52	Volume Not Met
Breath Volume		0.304	06:52	
Air Blank		0.000	06:52	
Control Test		0.077	06:53	
Air Blank		0.000	06:53	
Diagnostics Check		OK	06:53	

Data Download Date/Time: 03/31/2016 22:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PUSEY, ANTHONY, C
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: WURST, JOSHUA
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 02/19/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1752856
DRY GAS EXPIRY DATE: 10/13/2017
VIOLATION CODE: DUI
ARREST TIME: 00:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: ALDRIDGE

DATE/TIME OF TEST: 03/01/2016 01:26:51

Diagnostics Check		OK	01:20
Air Blank		0.000	01:21
Control Test		0.080	01:21
Air Blank		0.000	01:21
Subject Sample	# 1	0.188	01:22
Breath Volume		2.210	01:22
Air Blank		0.000	01:22
Air Blank		0.000	01:24
Subject Sample	# 2	0.192	01:25
Breath Volume		2.007	01:25
Air Blank		0.000	01:25
Control Test		0.080	01:26
Air Blank		0.000	01:26
Diagnostics Check		OK	01:26

Data Download Date/Time: 03/28/2016 10:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAYFIELD, MATTHEW, E
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: MAYFIELD, MATTHEW
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 02/19/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1752856
DRY GAS EXPIRY DATE: 10/13/2017
VIOLATION CODE: DUI
ARREST TIME: 22:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: GOODSON

DATE/TIME OF TEST: 03/04/2016 23:10:19

Diagnostics Check	OK	23:04
Air Blank	0.000	23:04
Control Test	0.081	23:05
Air Blank	0.000	23:05
Subject Sample # 1	0.000	23:06
Breath Volume	3.421	23:06
Air Blank	0.000	23:06
Air Blank	0.000	23:08
Subject Sample # 2	0.000	23:08
Breath Volume	3.726	23:08
Air Blank	0.000	23:09
Control Test	0.080	23:09
Air Blank	0.000	23:10
Diagnostics Check	OK	23:10

Data Download Date/Time: 03/28/2016 10:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANCE,CHARLES,R
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: LANCE,CHARLES
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 02/19/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1752856
DRY GAS EXPIRY DATE: 10/13/2017
VIOLATION CODE: DUI
ARREST TIME: 18:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:35

SUBJECT'S LAST NAME: CHOWBAY

DATE/TIME OF TEST: 03/05/2016 19:06:12

Diagnostics Check	OK	18:59
Air Blank	0.000	19:00
Control Test	0.079	19:00
Air Blank	0.000	19:00
Subject Sample # 1	0.177	19:01
Breath Volume	4.328	19:01
Air Blank	0.000	19:02
Air Blank	0.000	19:04
Subject Sample # 2	0.169	19:04
Breath Volume	2.257	19:04
Air Blank	0.000	19:05
Control Test	0.079	19:05
Air Blank	0.000	19:06
Diagnostics Check	OK	19:06

Data Download Date/Time: 03/28/2016 10:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PELFREY,PAUL,D
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: PELFREY,PAUL
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 02/19/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1752856
DRY GAS EXPIRY DATE: 10/13/2017
VIOLATION CODE: DUI
ARREST TIME: 04:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:15

SUBJECT'S LAST NAME: CHAVEZ

DATE/TIME OF TEST: 03/10/2016 04:56:31

Diagnostics Check	OK	04:47
Air Blank	0.000	04:48
Control Test	0.080	04:48
Air Blank	0.000	04:49
Subject Sample # 1	0.221	04:51
Breath Volume	1.632	04:51
Air Blank	0.000	04:52
Air Blank	0.000	04:54
Subject Sample # 2	0.219	04:55
Breath Volume	2.328	04:55
Air Blank	0.000	04:55
Control Test	0.079	04:55
Air Blank	0.000	04:56
Diagnostics Check	OK	04:56

Data Download Date/Time: 03/28/2016 10:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAYFIELD, MATTHEW, E
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: MAYFIELD, MATTHEW
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 02/19/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1752856
DRY GAS EXPIRY DATE: 10/13/2017
VIOLATION CODE: DUI
ARREST TIME: 04:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:40

SUBJECT'S LAST NAME: REAVIS

DATE/TIME OF TEST: 03/10/2016 05:09:13

Diagnostics Check	OK	05:03
Air Blank	0.000	05:03
Control Test	0.079	05:03
Air Blank	0.000	05:04
Subject Sample # 1	0.087	05:04
Breath Volume	1.597	05:04
Air Blank	0.000	05:05
Air Blank	0.000	05:07
Subject Sample # 2	0.084	05:07
Breath Volume	1.335	05:07
Air Blank	0.000	05:08
Control Test	0.079	05:08
Air Blank	0.000	05:09
Diagnostics Check	OK	05:09

Data Download Date/Time: 03/28/2016 10:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAYFIELD, MATTHEW, E
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: JORDAN, CLAYTON
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 02/19/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1752856
DRY GAS EXPIRY DATE: 10/13/2017
VIOLATION CODE: DUI
ARREST TIME: 20:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:25

SUBJECT'S LAST NAME: CARTER

DATE/TIME OF TEST: 03/10/2016 20:52:06

Diagnostics Check	OK	20:45
Air Blank	0.000	20:46
Control Test	0.079	20:46
Air Blank	0.000	20:47
Subject Sample # 1	0.173	20:47
Breath Volume	2.000	20:47
Air Blank	0.000	20:48
Air Blank	0.000	20:50
Subject Sample # 2	0.171	20:50
Breath Volume	2.058	20:50
Air Blank	0.000	20:51
Control Test	0.079	20:51
Air Blank	0.000	20:52
Diagnostics Check	OK	20:52

Data Download Date/Time: 03/28/2016 10:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PUSEY, ANTHONY, C
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: MOSS, ROBERT
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 02/19/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1752856
DRY GAS EXPIRY DATE: 10/13/2017
VIOLATION CODE: DUI
ARREST TIME: 21:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:05

SUBJECT'S LAST NAME: NIXON

DATE/TIME OF TEST: 03/10/2016 21:40:56

Diagnostics Check	OK	21:34
Air Blank	0.000	21:34
Control Test	0.079	21:35
Air Blank	0.000	21:35
Subject Sample # 1	0.091	21:36
Breath Volume	1.914	21:36
Air Blank	0.000	21:36
Air Blank	0.000	21:38
Subject Sample # 2	0.091	21:39
Breath Volume	1.503	21:39
Air Blank	0.000	21:40
Control Test	0.080	21:40
Air Blank	0.000	21:40
Diagnostics Check	OK	21:40

Data Download Date/Time: 03/28/2016 10:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAYFIELD, MATTHEW, E
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: MAYFIELD, MATTHEW
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 02/19/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1752856
DRY GAS EXPIRY DATE: 10/13/2017
VIOLATION CODE: DUI
ARREST TIME: 02:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: WADY ALANAZI

DATE/TIME OF TEST: 03/11/2016 03:04:23

Diagnostics Check	OK	02:57
Air Blank	0.000	02:57
Control Test	0.080	02:57
Air Blank	0.000	02:58
Subject Sample # 1	0.091	02:59
Breath Volume	2.308	02:59
Air Blank	0.000	03:00
Air Blank	0.000	03:02
Subject Sample # 2	0.089	03:02
Breath Volume	2.132	03:02
Air Blank	0.000	03:03
Control Test	0.080	03:03
Air Blank	0.000	03:04
Diagnostics Check	OK	03:04

Data Download Date/Time: 03/28/2016 10:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TOMLINSON,NICHOLAS,L
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: BROOKS,BLAKE
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 02/19/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1752856
DRY GAS EXPIRY DATE: 10/13/2017
VIOLATION CODE: DUI
ARREST TIME: 16:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:40

SUBJECT'S LAST NAME: BRUNING

DATE/TIME OF TEST: 03/11/2016 17:05:53

Diagnostics Check	OK	17:02	
Air Blank	0.000	17:03	
Control Test	0.080	17:03	
Air Blank	0.000	17:04	
Subject Sample	# 1	17:04	Subject Test Refused
Breath Volume	0.000	17:04	
Air Blank	0.000	17:05	
Control Test	0.080	17:05	
Air Blank	0.000	17:05	
Diagnostics Check	OK	17:05	

Data Download Date/Time: 03/28/2016 10:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,GARY,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: MILLER,GARY
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 02/19/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1752856
DRY GAS EXPIRY DATE: 10/13/2017
VIOLATION CODE: DUI
ARREST TIME: 20:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:25

SUBJECT'S LAST NAME: ORT

DATE/TIME OF TEST: 03/11/2016 21:57:09

Diagnostics Check	OK	21:50
Air Blank	0.000	21:51
Control Test	0.079	21:51
Air Blank	0.000	21:52
Subject Sample # 1	0.090	21:52
Breath Volume	2.113	21:52
Air Blank	0.000	21:53
Air Blank	0.000	21:55
Subject Sample # 2	0.083	21:55
Breath Volume	1.835	21:55
Air Blank	0.000	21:56
Control Test	0.079	21:56
Air Blank	0.000	21:57
Diagnostics Check	OK	21:57

Data Download Date/Time: 03/28/2016 10:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PELFREY,PAUL,D
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: PELFREY,PAUL
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 02/19/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1752856
DRY GAS EXPIRY DATE: 10/13/2017
VIOLATION CODE: DUI
ARREST TIME: 04:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:01

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 03/13/2016 05:16:42

Diagnostics Check	OK	05:09
Air Blank	0.000	05:10
Control Test	0.079	05:10
Air Blank	0.000	05:11
Subject Sample # 1	0.112	05:11
Breath Volume	3.902	05:11
Air Blank	0.000	05:12
Air Blank	0.000	05:14
Subject Sample # 2	0.099	05:15
Breath Volume	1.375	05:15
Air Blank	0.000	05:15
Control Test	0.079	05:16
Air Blank	0.000	05:16
Diagnostics Check	OK	05:16

Data Download Date/Time: 03/28/2016 10:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANCE, CHARLES, R
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: ROBERTS, BARRY
ARRESTING OFFICER AGENCY: Springfield PD
LAST AGENCY INSPECTION: 02/19/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1752856
DRY GAS EXPIRY DATE: 10/13/2017
VIOLATION CODE: DUI
ARREST TIME: 22:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: ZUNIGA

DATE/TIME OF TEST: 03/13/2016 22:48:09

Diagnostics Check	OK	22:40
Air Blank	0.000	22:41
Control Test	0.079	22:41
Air Blank	0.000	22:42
Subject Sample # 1	0.164	22:43
Breath Volume	2.300	22:43
Air Blank	0.000	22:44
Air Blank	0.000	22:46
Subject Sample # 2	0.174	22:46
Breath Volume	2.957	22:46
Air Blank	0.000	22:47
Control Test	0.077	22:47
Air Blank	0.000	22:48
Diagnostics Check	OK	22:48

Data Download Date/Time: 03/28/2016 10:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PUSEY, ANTHONY, C
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: MOSS, ROBERT
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 02/19/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1752856
DRY GAS EXPIRY DATE: 10/13/2017
VIOLATION CODE: DUI
ARREST TIME: 00:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:33

SUBJECT'S LAST NAME: MEUNIER

DATE/TIME OF TEST: 03/15/2016 01:13:07

Diagnostics Check		OK	00:57	
Air Blank		0.000	00:57	
Control Test		0.080	00:58	
Air Blank		0.000	00:58	
Subject Sample	# 1	0.251	01:01	Volume Not Met
Breath Volume		0.882	01:01	
Air Blank		0.000	01:02	
Air Blank		0.000	01:04	
Subject Sample	# 2	0.249	01:06	
Breath Volume		1.769	01:06	
Air Blank		0.000	01:07	
Air Blank		0.000	01:08	
Subject Sample	# 3	0.261	01:11	
Breath Volume		2.187	01:11	
Air Blank		0.000	01:12	
Control Test		0.077	01:12	
Air Blank		0.000	01:13	
Diagnostics Check		OK	01:13	

Data Download Date/Time: 03/28/2016 10:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PETTYS,JOEY,A
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: MILLER,GARY
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 02/19/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1752856
DRY GAS EXPIRY DATE: 10/13/2017
VIOLATION CODE: DUI
ARREST TIME: 21:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: COMSTOCK

DATE/TIME OF TEST: 03/15/2016 22:41:24

Diagnostics Check	OK	22:29	
Air Blank	0.000	22:30	
Control Test	0.078	22:30	
Air Blank	0.000	22:31	
Subject Sample	# 1 0.067	22:31	
Breath Volume	1.816	22:31	
Air Blank	0.000	22:32	
Air Blank	0.000	22:34	
Subject Sample	# 2 0.059	22:37	Volume Not Met
Breath Volume	0.277	22:37	
Air Blank	0.000	22:37	
Air Blank	0.000	22:39	
Subject Sample	# 3	22:40	Subject Test Refused
Breath Volume	0.000	22:40	
Air Blank	0.000	22:40	
Control Test	0.077	22:40	
Air Blank	0.000	22:41	
Diagnostics Check	OK	22:41	

Data Download Date/Time: 03/28/2016 10:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,CASEY,L
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: MOSS,BOB
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 02/19/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1752856
DRY GAS EXPIRY DATE: 10/13/2017
VIOLATION CODE: DUI
ARREST TIME: 03:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: ALSHEHRI

DATE/TIME OF TEST: 03/16/2016 03:38:12

Diagnostics Check		OK	03:31
Air Blank		0.000	03:32
Control Test		0.079	03:32
Air Blank		0.000	03:32
Subject Sample	# 1	0.254	03:33
Breath Volume		2.878	03:33
Air Blank		0.000	03:34
Air Blank		0.000	03:36
Subject Sample	# 2	0.261	03:36
Breath Volume		2.289	03:36
Air Blank		0.000	03:37
Control Test		0.079	03:37
Air Blank		0.000	03:38
Diagnostics Check		OK	03:38

Data Download Date/Time: 03/28/2016 10:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PELFREY,PAUL,D
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: PELFREY,PAUL
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 02/19/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1752856
DRY GAS EXPIRY DATE: 10/13/2017
VIOLATION CODE: DUI
ARREST TIME: 20:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:07

SUBJECT'S LAST NAME: HEFFEL

DATE/TIME OF TEST: 03/16/2016 21:42:25

Diagnostics Check		OK	21:30	
Air Blank		0.000	21:31	
Control Test		0.079	21:31	
Air Blank		0.000	21:32	
Subject Sample	# 1	0.115	21:35	Volume Not Met
Breath Volume		0.738	21:35	
Air Blank		0.000	21:35	
Air Blank		0.000	21:37	
Subject Sample	# 2	0.122	21:40	Volume Not Met
Breath Volume		0.316	21:40	
Air Blank		0.000	21:41	
Control Test		0.078	21:41	
Air Blank		0.000	21:42	
Diagnostics Check		OK	21:42	

Data Download Date/Time: 03/28/2016 10:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANCE,CHARLES,R
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: LANCE,CHARLES
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 02/19/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1752856
DRY GAS EXPIRY DATE: 10/13/2017
VIOLATION CODE: DUI
ARREST TIME: 23:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:58

SUBJECT'S LAST NAME: FORD

DATE/TIME OF TEST: 03/17/2016 00:25:04

Diagnostics Check		OK	00:18
Air Blank		0.000	00:18
Control Test		0.078	00:19
Air Blank		0.000	00:19
Subject Sample	# 1	0.211	00:20
Breath Volume		2.421	00:20
Air Blank		0.000	00:21
Air Blank		0.000	00:22
Subject Sample	# 2	0.202	00:23
Breath Volume		1.273	00:23
Air Blank		0.000	00:24
Control Test		0.077	00:24
Air Blank		0.000	00:24
Diagnostics Check		OK	00:25

Data Download Date/Time: 03/28/2016 10:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAYFIELD, MATTHEW, E
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: MAYFIELD, MATTHEW
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 02/19/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1752856
DRY GAS EXPIRY DATE: 10/13/2017
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: NOLAND

DATE/TIME OF TEST: 03/17/2016 02:27:20

Diagnostics Check	OK	02:21
Air Blank	0.000	02:21
Control Test	0.079	02:21
Air Blank	0.000	02:22
Subject Sample # 1	0.100	02:22
Breath Volume	1.375	02:22
Air Blank	0.000	02:23
Air Blank	0.000	02:25
Subject Sample # 2	0.093	02:25
Breath Volume	2.343	02:25
Air Blank	0.000	02:26
Control Test	0.078	02:26
Air Blank	0.000	02:27
Diagnostics Check	OK	02:27

Data Download Date/Time: 03/28/2016 10:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PELFREY,PAUL,D
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: RHODES,JASON
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 02/19/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1752856
DRY GAS EXPIRY DATE: 10/13/2017
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: DUDLEY

DATE/TIME OF TEST: 03/20/2016 01:22:14

Diagnostics Check		OK	01:10	
Air Blank		0.000	01:11	
Control Test		0.079	01:11	
Air Blank		0.000	01:11	
Subject Sample	# 1	0.140	01:15	Volume Not Met
Breath Volume		0.304	01:15	
Air Blank		0.000	01:15	
Air Blank		0.000	01:17	
Subject Sample	# 2	0.157	01:20	Volume Not Met
Breath Volume		0.066	01:20	
Air Blank		0.000	01:21	
Control Test		0.079	01:21	
Air Blank		0.000	01:22	
Diagnostics Check		OK	01:22	

Data Download Date/Time: 03/28/2016 10:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PELFREY,PAUL,D
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: RHODES,JASON
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 02/19/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1752856
DRY GAS EXPIRY DATE: 10/13/2017
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: DUDLEY

DATE/TIME OF TEST: 03/20/2016 01:39:53

Diagnostics Check		OK	01:28	
Air Blank		0.000	01:28	
Control Test		0.077	01:29	
Air Blank		0.000	01:29	
Subject Sample	# 1	0.148	01:32	Volume Not Met
Breath Volume		0.210	01:32	
Air Blank		0.000	01:33	
Air Blank		0.000	01:35	
Subject Sample	# 2	0.155	01:38	Volume Not Met
Breath Volume		0.312	01:38	
Air Blank		0.000	01:38	
Control Test		0.077	01:39	
Air Blank		0.000	01:39	
Diagnostics Check		OK	01:39	

Data Download Date/Time: 03/28/2016 10:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,CASEY,L
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: MOSS,ROBERT
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 02/19/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1752856
DRY GAS EXPIRY DATE: 10/13/2017
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: FAULKNER

DATE/TIME OF TEST: 03/23/2016 03:01:05

Diagnostics Check	OK	02:54
Air Blank	0.000	02:54
Control Test	0.081	02:55
Air Blank	0.000	02:55
Subject Sample # 1	0.214	02:56
Breath Volume	2.417	02:56
Air Blank	0.000	02:56
Air Blank	0.000	02:58
Subject Sample # 2	0.213	02:59
Breath Volume	3.140	02:59
Air Blank	0.000	03:00
Control Test	0.081	03:00
Air Blank	0.000	03:01
Diagnostics Check	OK	03:01

Data Download Date/Time: 03/28/2016 10:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,CASEY,L
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: WURST,JOSHUA
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 02/19/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1752856
DRY GAS EXPIRY DATE: 10/13/2017
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: DUNN

DATE/TIME OF TEST: 03/26/2016 03:25:14

Diagnostics Check	OK	03:18
Air Blank	0.000	03:19
Control Test	0.078	03:19
Air Blank	0.000	03:19
Subject Sample # 1	0.270	03:20
Breath Volume	4.117	03:20
Air Blank	0.000	03:21
Air Blank	0.000	03:23
Subject Sample # 2	0.268	03:23
Breath Volume	3.507	03:23
Air Blank	0.000	03:24
Control Test	0.077	03:24
Air Blank	0.000	03:25
Diagnostics Check	OK	03:25

Data Download Date/Time: 03/28/2016 10:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PELFREY,PAUL,D
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: MOSS,ROBERT
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 02/19/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1752856
DRY GAS EXPIRY DATE: 10/13/2017
VIOLATION CODE: DUI
ARREST TIME: 21:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: TIDWELL

DATE/TIME OF TEST: 03/26/2016 22:46:39

Diagnostics Check	OK	22:40
Air Blank	0.000	22:40
Control Test	0.080	22:40
Air Blank	0.000	22:41
Subject Sample # 1	0.000	22:42
Breath Volume	3.296	22:42
Air Blank	0.000	22:42
Air Blank	0.000	22:44
Subject Sample # 2	0.000	22:45
Breath Volume	2.949	22:45
Air Blank	0.000	22:45
Control Test	0.079	22:46
Air Blank	0.000	22:46
Diagnostics Check	OK	22:46

Data Download Date/Time: 03/28/2016 10:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: LYNN HAVEN PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NUSBAUM,K,
BREATH TEST OPERATOR AGENCY: Lynn Haven PD
ARRESTING OFFICER: PAUGA,J
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 02/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29614080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 00:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:31

SUBJECT'S LAST NAME: TURNER

DATE/TIME OF TEST: 03/06/2016 01:29:47

Diagnostics Check	OK	01:23
Air Blank	0.000	01:23
Control Test	0.078	01:23
Air Blank	0.000	01:24
Subject Sample # 1	0.176	01:25
Breath Volume	1.187	01:25
Air Blank	0.000	01:25
Air Blank	0.000	01:27
Subject Sample # 2	0.161	01:28
Breath Volume	2.062	01:28
Air Blank	0.000	01:28
Control Test	0.077	01:29
Air Blank	0.000	01:29
Diagnostics Check	OK	01:29

Data Download Date/Time: 03/29/2016 15:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: LYNN HAVEN PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NUSBAUM,K,
BREATH TEST OPERATOR AGENCY: Lynn Haven PD
ARRESTING OFFICER: NUSBAUM,K
ARRESTING OFFICER AGENCY: Lynn Haven PD
LAST AGENCY INSPECTION: 02/29/2016

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 29614080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 20:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:49

SUBJECT'S LAST NAME: MCLENDON

DATE/TIME OF TEST: 03/20/2016 22:00:11

Diagnostics Check	OK	21:57	
Air Blank	0.000	21:57	
Control Test	0.077	21:58	
Air Blank	0.000	21:58	
Subject Sample	# 1	21:58	Subject Test Refused
Breath Volume	0.000	21:58	
Air Blank	0.000	21:59	
Control Test	0.078	21:59	
Air Blank	0.000	22:00	
Diagnostics Check	OK	22:00	

Data Download Date/Time: 03/29/2016 15:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: LYNN HAVEN PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NUSBAUM,K,
BREATH TEST OPERATOR AGENCY: Lynn Haven PD
ARRESTING OFFICER: NUSBAUM,K
ARRESTING OFFICER AGENCY: Lynn Haven PD
LAST AGENCY INSPECTION: 02/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29614080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 23:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: SPIVEY

DATE/TIME OF TEST: 03/21/2016 00:08:22

Diagnostics Check	OK	00:02
Air Blank	0.000	00:02
Control Test	0.078	00:02
Air Blank	0.000	00:03
Subject Sample # 1	0.000	00:03
Breath Volume	2.492	00:03
Air Blank	0.000	00:04
Air Blank	0.000	00:06
Subject Sample # 2	0.000	00:07
Breath Volume	1.628	00:07
Air Blank	0.000	00:07
Control Test	0.076	00:07
Air Blank	0.000	00:08
Diagnostics Check	OK	00:08

Data Download Date/Time: 03/29/2016 15:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CANNON,JAMES,J
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: JOHNSON,SARAH
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 02/17/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:54

SUBJECT'S LAST NAME: SUTHERIN

DATE/TIME OF TEST: 03/01/2016 01:22:44

Diagnostics Check	OK	01:16
Air Blank	0.000	01:16
Control Test	0.080	01:16
Air Blank	0.000	01:17
Subject Sample # 1	0.201	01:18
Breath Volume	2.152	01:18
Air Blank	0.000	01:18
Air Blank	0.000	01:20
Subject Sample # 2	0.190	01:21
Breath Volume	1.828	01:21
Air Blank	0.000	01:21
Control Test	0.080	01:22
Air Blank	0.000	01:22
Diagnostics Check	OK	01:22

Data Download Date/Time: 03/09/2016 09:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FESSEL,DANIEL,C
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: ROSALES,PATRICK
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 02/17/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:37

SUBJECT'S LAST NAME: FRIGGE

DATE/TIME OF TEST: 03/05/2016 01:08:51

Diagnostics Check		OK	01:02
Air Blank		0.000	01:03
Control Test		0.080	01:03
Air Blank		0.000	01:03
Subject Sample	# 1	0.086	01:04
Breath Volume		2.960	01:04
Air Blank		0.000	01:05
Air Blank		0.000	01:06
Subject Sample	# 2	0.080	01:07
Breath Volume		2.093	01:07
Air Blank		0.000	01:07
Control Test		0.080	01:08
Air Blank		0.000	01:08
Diagnostics Check		OK	01:08

Data Download Date/Time: 03/09/2016 09:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CANNON,JAMES,J
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: TUCKER,HAILEY
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 02/17/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: MERCK

DATE/TIME OF TEST: 03/06/2016 23:55:33

Diagnostics Check	OK	23:47
Air Blank	0.000	23:48
Control Test	0.081	23:48
Air Blank	0.000	23:48
Subject Sample # 1	0.216	23:51
Breath Volume	1.386	23:51
Air Blank	0.000	23:51
Air Blank	0.000	23:53
Subject Sample # 2	0.224	23:54
Breath Volume	1.742	23:54
Air Blank	0.000	23:54
Control Test	0.080	23:54
Air Blank	0.000	23:55
Diagnostics Check	OK	23:55

Data Download Date/Time: 03/09/2016 09:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZARRILLO,N,
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: HANDLEY,A
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 02/17/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:28

SUBJECT'S LAST NAME: EDWARDS

DATE/TIME OF TEST: 03/07/2016 21:54:13

Diagnostics Check	OK	21:48	
Air Blank	0.000	21:48	
Control Test	0.081	21:49	
Air Blank	0.000	21:49	
Subject Sample	# 1 0.000	21:50	
Breath Volume	3.441	21:50	
Air Blank	0.000	21:50	
Air Blank	0.000	21:52	
Subject Sample	# 2 0.000	21:53	
Breath Volume	3.070	21:53	
Air Blank	0.000	21:53	
Control Test	0.000	21:53	RFI Detect
Air Blank	0.000	21:54	

Data Download Date/Time: 03/09/2016 09:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUSKEVICH,CURT,E
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: SUSKEVICH,CURT
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 03/09/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:53

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 03/10/2016 03:19:33

Diagnostics Check		OK	03:13
Air Blank		0.000	03:14
Control Test		0.080	03:14
Air Blank		0.000	03:14
Subject Sample	# 1	0.050	03:15
Breath Volume		2.406	03:15
Air Blank		0.000	03:15
Air Blank		0.000	03:17
Subject Sample	# 2	0.049	03:18
Breath Volume		2.898	03:18
Air Blank		0.000	03:18
Control Test		0.080	03:18
Air Blank		0.000	03:19
Diagnostics Check		OK	03:19

Data Download Date/Time: 04/13/2016 08:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAUSER,SGT.,H
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: CINTRON,M
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 03/09/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: REYNOLDS

DATE/TIME OF TEST: 03/11/2016 02:53:21

Diagnostics Check	OK	02:46
Air Blank	0.000	02:47
Control Test	0.080	02:47
Air Blank	0.000	02:47
Subject Sample # 1	0.152	02:48
Breath Volume	1.109	02:48
Air Blank	0.000	02:49
Air Blank	0.000	02:51
Subject Sample # 2	0.172	02:51
Breath Volume	1.472	02:51
Air Blank	0.000	02:52
Control Test	0.080	02:52
Air Blank	0.000	02:53
Diagnostics Check	OK	02:53

Data Download Date/Time: 04/13/2016 08:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'GRADY,PATRICK,
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: O'GRADY,PATRICK
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 03/09/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:53

SUBJECT'S LAST NAME: UDAS

DATE/TIME OF TEST: 03/12/2016 01:30:49

Diagnostics Check	OK	01:24
Air Blank	0.000	01:24
Control Test	0.080	01:25
Air Blank	0.000	01:25
Subject Sample # 1	0.120	01:25
Breath Volume	4.230	01:25
Air Blank	0.000	01:26
Air Blank	0.000	01:28
Subject Sample # 2	0.119	01:29
Breath Volume	2.957	01:29
Air Blank	0.000	01:29
Control Test	0.081	01:30
Air Blank	0.000	01:30
Diagnostics Check	OK	01:30

Data Download Date/Time: 04/13/2016 08:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARKLEY,JEFFREY,G
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: BARKLEY,JEFFREY
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 03/09/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:06

SUBJECT'S LAST NAME: MILLS

DATE/TIME OF TEST: 03/13/2016 23:32:45

Diagnostics Check	OK	23:26
Air Blank	0.000	23:27
Control Test	0.080	23:27
Air Blank	0.000	23:27
Subject Sample # 1	0.216	23:28
Breath Volume	3.500	23:28
Air Blank	0.000	23:28
Air Blank	0.000	23:30
Subject Sample # 2	0.212	23:31
Breath Volume	2.886	23:31
Air Blank	0.000	23:31
Control Test	0.079	23:32
Air Blank	0.000	23:32
Diagnostics Check	OK	23:32

Data Download Date/Time: 04/13/2016 08:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUFFY, JONATHAN,
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: OGRADY, P
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 03/09/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 17:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:42

SUBJECT'S LAST NAME: SIMON

DATE/TIME OF TEST: 03/15/2016 18:06:12

Diagnostics Check	OK	18:03	
Air Blank	0.000	18:03	
Control Test	0.079	18:04	
Air Blank	0.000	18:04	
Subject Sample	# 1 0.117	18:05	
Breath Volume	3.167	18:05	
Air Blank		18:05	RFI Detect
Air Blank		18:06	RFI Detect
Air Blank		18:06	RFI Detect

Data Download Date/Time: 04/13/2016 08:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUFFY, JONATHAN,
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: OGRADY, P
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 03/09/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 17:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:42

SUBJECT'S LAST NAME: SIMON

DATE/TIME OF TEST: 03/15/2016 18:14:20

Diagnostics Check	OK	18:08
Air Blank	0.000	18:08
Control Test	0.080	18:08
Air Blank	0.000	18:09
Subject Sample # 1	0.113	18:09
Breath Volume	3.296	18:09
Air Blank	0.000	18:10
Air Blank	0.000	18:12
Subject Sample # 2	0.108	18:12
Breath Volume	2.621	18:12
Air Blank	0.000	18:13
Control Test	0.080	18:13
Air Blank	0.000	18:14
Diagnostics Check	OK	18:14

Data Download Date/Time: 04/13/2016 08:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FESSEL,DANIEL,C
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: THOMAS,DANIEL
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 03/09/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 19:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:56

SUBJECT'S LAST NAME: FRENCH

DATE/TIME OF TEST: 03/15/2016 20:34:50

Diagnostics Check	OK	20:28
Air Blank	0.000	20:29
Control Test	0.080	20:29
Air Blank	0.000	20:30
Subject Sample # 1	0.245	20:30
Breath Volume	2.628	20:30
Air Blank	0.000	20:31
Air Blank	0.000	20:32
Subject Sample # 2	0.252	20:33
Breath Volume	2.437	20:33
Air Blank	0.000	20:33
Control Test	0.080	20:34
Air Blank	0.000	20:34
Diagnostics Check	OK	20:34

Data Download Date/Time: 04/13/2016 08:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARSON,JAMES,A
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: WALLACE,ANNA
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 03/09/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:49

SUBJECT'S LAST NAME: HESS

DATE/TIME OF TEST: 03/17/2016 00:19:04

Diagnostics Check	OK	00:11
Air Blank	0.000	00:11
Control Test	0.080	00:12
Air Blank	0.000	00:12
Subject Sample # 1	0.200	00:14
Breath Volume	1.886	00:14
Air Blank	0.000	00:14
Air Blank	0.000	00:16
Subject Sample # 2	0.193	00:17
Breath Volume	1.468	00:17
Air Blank	0.000	00:18
Control Test	0.079	00:18
Air Blank	0.000	00:18
Diagnostics Check	OK	00:19

Data Download Date/Time: 04/13/2016 08:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLS,CHARLES,S
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MILLS,CHARLES
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 03/09/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:18

SUBJECT'S LAST NAME: ORTEGA

DATE/TIME OF TEST: 03/18/2016 00:50:07

Diagnostics Check	OK	00:42
Air Blank	0.000	00:42
Control Test	0.079	00:43
Air Blank	0.000	00:43
Subject Sample # 1	0.158	00:44
Breath Volume	1.210	00:44
Air Blank	0.000	00:44
Air Blank	0.000	00:46
Subject Sample # 2	0.146	00:48
Breath Volume	1.328	00:48
Air Blank	0.000	00:49
Control Test	0.080	00:49
Air Blank	0.000	00:50
Diagnostics Check	OK	00:50

Data Download Date/Time: 04/13/2016 08:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAUSER,SGT,H
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: DURHAM,K
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 03/09/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:33

SUBJECT'S LAST NAME: HUDSON

DATE/TIME OF TEST: 03/18/2016 22:11:48

Diagnostics Check	OK	22:02
Air Blank	0.000	22:02
Control Test	0.081	22:03
Air Blank	0.000	22:03
Subject Sample # 1	0.123	22:05
Breath Volume	1.343	22:05
Air Blank	0.000	22:06
Air Blank	0.000	22:08
Subject Sample # 2	0.115	22:10
Breath Volume	2.183	22:10
Air Blank	0.000	22:10
Control Test	0.081	22:11
Air Blank	0.000	22:11
Diagnostics Check	OK	22:11

Data Download Date/Time: 04/13/2016 08:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEMOS,MATHEW,T
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: LEMOS,MATHEW
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 03/09/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 18:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:45

SUBJECT'S LAST NAME: FRIEDEL

DATE/TIME OF TEST: 03/19/2016 19:14:04

Diagnostics Check	OK	19:07
Air Blank	0.000	19:08
Control Test	0.080	19:08
Air Blank	0.000	19:09
Subject Sample # 1	0.316	19:09
Breath Volume	1.726	19:09
Air Blank	0.000	19:10
Air Blank	0.000	19:12
Subject Sample # 2	0.324	19:12
Breath Volume	1.953	19:12
Air Blank	0.000	19:13
Control Test	0.080	19:13
Air Blank	0.000	19:13
Diagnostics Check	OK	19:14

Data Download Date/Time: 04/13/2016 08:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FESSEL,DANIEL,C
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: HAMILTON-REYES,CORTNEY
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 03/09/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:04

SUBJECT'S LAST NAME: DOYON

DATE/TIME OF TEST: 03/19/2016 22:34:00

Diagnostics Check	OK	22:27
Air Blank	0.000	22:28
Control Test	0.081	22:28
Air Blank	0.000	22:28
Subject Sample # 1	0.226	22:29
Breath Volume	1.925	22:29
Air Blank	0.000	22:30
Air Blank	0.000	22:31
Subject Sample # 2	0.217	22:32
Breath Volume	1.761	22:32
Air Blank	0.000	22:33
Control Test	0.080	22:33
Air Blank	0.000	22:33
Diagnostics Check	OK	22:34

Data Download Date/Time: 04/13/2016 08:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARKLEY,JEFFREY,G
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: CARVALHO,ASHLEY
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 03/09/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:00

SUBJECT'S LAST NAME: COMBS

DATE/TIME OF TEST: 03/26/2016 05:29:24

Diagnostics Check		OK	05:20
Air Blank		0.000	05:21
Control Test		0.081	05:21
Air Blank		0.000	05:22
Subject Sample	# 1	0.250	05:24
Breath Volume		2.417	05:24
Air Blank		0.000	05:25
Air Blank		0.000	05:26
Subject Sample	# 2	0.247	05:27
Breath Volume		3.328	05:27
Air Blank		0.000	05:28
Control Test		0.079	05:28
Air Blank		0.000	05:29
Diagnostics Check		OK	05:29

Data Download Date/Time: 04/13/2016 08:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZARRILLO,N,
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: HANDLEY,A
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 03/09/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:06

SUBJECT'S LAST NAME: OLIVA

DATE/TIME OF TEST: 03/27/2016 02:32:17

Diagnostics Check	OK	02:26
Air Blank	0.000	02:26
Control Test	0.080	02:27
Air Blank	0.000	02:27
Subject Sample # 1	0.056	02:27
Breath Volume	2.906	02:27
Air Blank	0.000	02:28
Air Blank	0.000	02:30
Subject Sample # 2	0.054	02:30
Breath Volume	2.847	02:30
Air Blank	0.000	02:31
Control Test	0.080	02:31
Air Blank	0.000	02:32
Diagnostics Check	OK	02:32

Data Download Date/Time: 04/13/2016 08:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CANNON,JAMES,J
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: HAMILTON-REYES,CORTNEY
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 03/09/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:42

SUBJECT'S LAST NAME: WAINEY

DATE/TIME OF TEST: 03/29/2016 02:13:10

Diagnostics Check	OK	02:05
Air Blank	0.000	02:06
Control Test	0.080	02:06
Air Blank	0.000	02:07
Subject Sample # 1	0.183	02:07
Breath Volume	1.292	02:07
Air Blank	0.000	02:08
Air Blank	0.000	02:10
Subject Sample # 2	0.184	02:11
Breath Volume	2.058	02:11
Air Blank	0.000	02:12
Control Test	0.080	02:12
Air Blank	0.000	02:13
Diagnostics Check	OK	02:13

Data Download Date/Time: 04/13/2016 08:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARKLEY,JEFFREY,G
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: BARKLEY,JEFFREY
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 03/09/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: ALLESSIO

DATE/TIME OF TEST: 03/31/2016 00:07:21

Diagnostics Check	OK	00:01
Air Blank	0.000	00:01
Control Test	0.080	00:01
Air Blank	0.000	00:02
Subject Sample # 1	0.000	00:02
Breath Volume	3.695	00:02
Air Blank	0.000	00:03
Air Blank	0.000	00:05
Subject Sample # 2	0.000	00:06
Breath Volume	3.492	00:06
Air Blank	0.000	00:06
Control Test	0.081	00:06
Air Blank	0.000	00:07
Diagnostics Check	OK	00:07

Data Download Date/Time: 04/13/2016 08:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MINGS,DEPUTY G.,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: MINGS,DEPUTY
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 02/23/2016

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 28814080A3
DRY GAS EXPIRY DATE: 11/05/2016
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 03/09/2016 03:12:10

Diagnostics Check	OK	03:05
Air Blank	0.000	03:06
Control Test	0.081	03:06
Air Blank	0.000	03:06
Subject Sample # 1	0.117	03:07
Breath Volume	1.367	03:07
Air Blank	0.000	03:08
Air Blank	0.000	03:09
Subject Sample # 2	0.114	03:10
Breath Volume	1.988	03:10
Air Blank	0.000	03:11
Control Test	0.078	03:11
Air Blank	0.000	03:12
Diagnostics Check	OK	03:12

Data Download Date/Time: 03/25/2016 11:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MINGS,DEPUTY,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: MINGS,DEPUTY G.
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 02/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28814080A3
DRY GAS EXPIRY DATE: 11/05/2016
VIOLATION CODE: DUI
ARREST TIME: 20:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:42

SUBJECT'S LAST NAME: KVAMME

DATE/TIME OF TEST: 03/15/2016 20:41:43

Diagnostics Check	OK	20:35
Air Blank	0.000	20:36
Control Test	0.080	20:36
Air Blank	0.000	20:36
Subject Sample # 1	0.256	20:37
Breath Volume	1.367	20:37
Air Blank	0.000	20:37
Air Blank	0.000	20:39
Subject Sample # 2	0.260	20:40
Breath Volume	2.992	20:40
Air Blank	0.000	20:40
Control Test	0.079	20:41
Air Blank	0.000	20:41
Diagnostics Check	OK	20:41

Data Download Date/Time: 03/25/2016 11:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MINGS,DEPUTY G.,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: MINGS,DEPUTY G.
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 02/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28814080A3
DRY GAS EXPIRY DATE: 11/05/2016
VIOLATION CODE: DUI
ARREST TIME: 02:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:38

SUBJECT'S LAST NAME: COOK

DATE/TIME OF TEST: 03/19/2016 03:04:42

Diagnostics Check	OK	02:58
Air Blank	0.000	02:58
Control Test	0.080	02:59
Air Blank	0.000	02:59
Subject Sample # 1	0.157	03:00
Breath Volume	2.093	03:00
Air Blank	0.000	03:00
Air Blank	0.000	03:02
Subject Sample # 2	0.151	03:03
Breath Volume	1.753	03:03
Air Blank	0.000	03:03
Control Test	0.080	03:04
Air Blank	0.000	03:04
Diagnostics Check	OK	03:04

Data Download Date/Time: 03/25/2016 11:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CROSS,B,Y
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: CROSS,B
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 02/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1816818
DRY GAS EXPIRY DATE: 02/13/2018
VIOLATION CODE: DUI
ARREST TIME: 09:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:50

SUBJECT'S LAST NAME: HOLT

DATE/TIME OF TEST: 03/01/2016 10:25:04

Diagnostics Check		OK	10:13	
Air Blank		0.000	10:13	
Control Test		0.081	10:14	
Air Blank		0.000	10:14	
Subject Sample	# 1	0.240	10:17	Volume Not Met
Breath Volume		0.335	10:17	
Air Blank		0.000	10:18	
Air Blank		0.000	10:20	
Subject Sample	# 2	0.307	10:23	Volume Not Met
Breath Volume		0.183	10:23	
Air Blank		0.000	10:24	
Control Test		0.078	10:24	
Air Blank		0.000	10:24	
Diagnostics Check		OK	10:25	

Data Download Date/Time: 03/10/2016 08:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CROSS,BRYAN,Y
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: CROSS,B
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 02/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1816818
DRY GAS EXPIRY DATE: 02/13/2018
VIOLATION CODE: DUI
ARREST TIME: 09:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:00

SUBJECT'S LAST NAME: HOLT

DATE/TIME OF TEST: 03/01/2016 10:40:12

Diagnostics Check		OK	10:28	
Air Blank		0.000	10:28	
Control Test		0.079	10:29	
Air Blank		0.000	10:29	
Subject Sample	# 1	0.272	10:32	Volume Not Met
Breath Volume		0.066	10:32	
Air Blank		0.000	10:33	
Air Blank		0.000	10:35	
Subject Sample	# 2	0.265	10:38	Volume Not Met
Breath Volume		0.203	10:38	
Air Blank		0.000	10:39	
Control Test		0.078	10:39	
Air Blank		0.000	10:40	
Diagnostics Check		OK	10:40	

Data Download Date/Time: 03/10/2016 08:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CROSS,B,Y
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: CROSS,B
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 02/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1816818
DRY GAS EXPIRY DATE: 02/13/2018
VIOLATION CODE: DUI
ARREST TIME: 10:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:15

SUBJECT'S LAST NAME: HOLT

DATE/TIME OF TEST: 03/01/2016 10:58:43

Diagnostics Check		OK	10:42	
Air Blank		0.000	10:43	
Control Test		0.080	10:43	
Air Blank		0.000	10:43	
Subject Sample	# 1	0.308	10:47	Volume Not Met
Breath Volume		0.421	10:47	
Air Blank		0.000	10:47	
Air Blank		0.000	10:49	
Subject Sample	# 2	0.286	10:51	
Breath Volume		1.507	10:51	
Air Blank		0.000	10:52	
Air Blank		0.000	10:53	
Subject Sample	# 3	0.303	10:57	Volume Not Met
Breath Volume		0.062	10:57	
Air Blank		0.000	10:57	
Control Test		0.077	10:58	
Air Blank		0.000	10:58	
Diagnostics Check		OK	10:58	

Data Download Date/Time: 03/10/2016 08:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CROSS,B,Y
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: CROSS,B
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 02/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1816818
DRY GAS EXPIRY DATE: 02/13/2018
VIOLATION CODE: DUI
ARREST TIME: 10:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:40

SUBJECT'S LAST NAME: HOLT

DATE/TIME OF TEST: 03/01/2016 11:08:41

Diagnostics Check		OK	11:01
Air Blank		0.000	11:01
Control Test		0.079	11:02
Air Blank		0.000	11:02
Subject Sample	# 1	0.274	11:04
Breath Volume		1.347	11:04
Air Blank		0.000	11:04
Air Blank		0.000	11:06
Subject Sample	# 2	0.285	11:07
Breath Volume		1.585	11:07
Air Blank		0.000	11:07
Control Test		0.078	11:08
Air Blank		0.000	11:08
Diagnostics Check		OK	11:08

Data Download Date/Time: 03/10/2016 08:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANCHEZ,JESUS,
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: TRUEBLOOD,RYAN
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 03/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1816818
DRY GAS EXPIRY DATE: 02/13/2018
VIOLATION CODE: DUI
ARREST TIME: 19:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: SANCHEZ

DATE/TIME OF TEST: 03/13/2016 20:28:26

Diagnostics Check	OK	20:25	
Air Blank	0.000	20:26	
Control Test	0.080	20:26	
Air Blank	0.000	20:26	
Subject Sample	# 1	20:27	Subject Test Refused
Breath Volume	0.000	20:27	
Air Blank	0.000	20:27	
Control Test	0.080	20:27	
Air Blank	0.000	20:28	
Diagnostics Check	OK	20:28	

Data Download Date/Time: 04/28/2016 09:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORBETA,JORGE,
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: TRUEBLOOD,RYAN
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 03/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1816818
DRY GAS EXPIRY DATE: 02/13/2018
VIOLATION CODE: DUI
ARREST TIME: 19:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: SANCHEZ

DATE/TIME OF TEST: 03/13/2016 20:37:34

Diagnostics Check	OK	20:34	
Air Blank	0.000	20:34	
Control Test	0.080	20:35	
Air Blank	0.000	20:35	
Subject Sample	# 1	20:36	Subject Test Refused
Breath Volume	0.000	20:36	
Air Blank	0.000	20:36	
Control Test	0.080	20:36	
Air Blank	0.000	20:37	
Diagnostics Check	OK	20:37	

Data Download Date/Time: 04/28/2016 09:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,GIL,A
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: GONZALEZ,GIL
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 03/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1816818
DRY GAS EXPIRY DATE: 02/13/2018
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:26

SUBJECT'S LAST NAME: PIETRYLO

DATE/TIME OF TEST: 03/19/2016 03:59:34

Diagnostics Check	OK	03:56	
Air Blank	0.000	03:57	
Control Test	0.080	03:57	
Air Blank	0.000	03:58	
Subject Sample	# 1	03:58	Subject Test Refused
Breath Volume	0.000	03:58	
Air Blank	0.000	03:58	
Control Test	0.081	03:58	
Air Blank	0.000	03:59	
Diagnostics Check	OK	03:59	

Data Download Date/Time: 04/28/2016 09:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRACEY,CHRISTINE,L
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: HODGES,T
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 03/10/2016

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1816818
DRY GAS EXPIRY DATE: 02/13/2018
VIOLATION CODE: DUI
ARREST TIME: 03:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:40

SUBJECT'S LAST NAME: PARE

DATE/TIME OF TEST: 03/24/2016 05:04:00

Diagnostics Check	OK	05:01	
Air Blank	0.000	05:01	
Control Test	0.080	05:01	
Air Blank	0.000	05:02	
Subject Sample	# 1	05:02	Subject Test Refused
Breath Volume	0.000	05:02	
Air Blank	0.000	05:03	
Control Test	0.080	05:03	
Air Blank	0.000	05:03	
Diagnostics Check	OK	05:04	

Data Download Date/Time: 04/28/2016 09:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLAUDY,MICHAEL,S
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: CLAUDY,MICHAEL
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 03/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1816818
DRY GAS EXPIRY DATE: 02/13/2018
VIOLATION CODE: DUI
ARREST TIME: 03:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:38

SUBJECT'S LAST NAME: FERRARO

DATE/TIME OF TEST: 03/27/2016 05:37:42

Diagnostics Check	OK	05:34	
Air Blank	0.000	05:35	
Control Test	0.081	05:35	
Air Blank	0.000	05:36	
Subject Sample	# 1	05:36	Subject Test Refused
Breath Volume	0.000	05:36	
Air Blank	0.000	05:36	
Control Test	0.081	05:37	
Air Blank	0.000	05:37	
Diagnostics Check	OK	05:37	

Data Download Date/Time: 04/28/2016 09:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VILLANUEVA, MAURICIO,
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: SANCHEZ, LUIS
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 03/10/2016

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1816818
DRY GAS EXPIRY DATE: 02/13/2018
VIOLATION CODE: DUI
ARREST TIME: 18:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:22

SUBJECT'S LAST NAME: ALCALA YERO

DATE/TIME OF TEST: 03/27/2016 19:45:53

Diagnostics Check	OK	19:42	
Air Blank	0.000	19:43	
Control Test	0.081	19:43	
Air Blank	0.000	19:43	
Subject Sample	# 1	19:44	Subject Test Refused
Breath Volume	0.000	19:44	
Air Blank	0.000	19:44	
Control Test	0.080	19:45	
Air Blank	0.000	19:45	
Diagnostics Check	OK	19:45	

Data Download Date/Time: 04/28/2016 09:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VILLANUEVA, MAURICIO,
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: WILKINS, MALACHI
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 03/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1816818
DRY GAS EXPIRY DATE: 02/13/2018
VIOLATION CODE: DUI
ARREST TIME: 18:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:30

SUBJECT'S LAST NAME: MOLINARO

DATE/TIME OF TEST: 03/27/2016 20:09:13

Diagnostics Check	OK	20:01
Air Blank	0.000	20:01
Control Test	0.081	20:01
Air Blank	0.000	20:02
Subject Sample # 1	0.162	20:03
Breath Volume	1.250	20:03
Air Blank	0.000	20:03
Air Blank	0.000	20:05
Subject Sample # 2	0.181	20:07
Breath Volume	2.566	20:07
Air Blank	0.000	20:08
Control Test	0.081	20:08
Air Blank	0.000	20:09
Diagnostics Check	OK	20:09

Data Download Date/Time: 04/28/2016 09:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,DEANNA,N
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: KERNS,CODY
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 03/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1816818
DRY GAS EXPIRY DATE: 02/13/2018
VIOLATION CODE: DUI
ARREST TIME: 02:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:49

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 03/31/2016 03:18:31

Diagnostics Check	OK	03:09
Air Blank	0.000	03:09
Control Test	0.081	03:10
Air Blank	0.000	03:10
Subject Sample # 1	0.269	03:11
Breath Volume	4.292	03:11
Air Blank	0.000	03:11
Air Blank	0.000	03:13
Subject Sample # 2	0.216	03:13
Breath Volume	1.605	03:13
Air Blank	0.000	03:14
Air Blank	0.000	03:16
Subject Sample # 3	0.267	03:17
Breath Volume	4.398	03:17
Air Blank	0.000	03:17
Control Test	0.081	03:17
Air Blank	0.000	03:18
Diagnostics Check	OK	03:18

Data Download Date/Time: 04/28/2016 09:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,DEANNA,N
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: SANCHEZ,LUIS
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 03/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1816818
DRY GAS EXPIRY DATE: 02/13/2018
VIOLATION CODE: DUI
ARREST TIME: 19:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 03/31/2016 21:23:28

Diagnostics Check	OK	21:20	
Air Blank	0.000	21:21	
Control Test	0.081	21:21	
Air Blank	0.000	21:21	
Subject Sample	# 1	21:22	Subject Test Refused
Breath Volume	0.000	21:22	
Air Blank	0.000	21:22	
Control Test	0.080	21:22	
Air Blank	0.000	21:23	
Diagnostics Check	OK	21:23	

Data Download Date/Time: 04/28/2016 09:19