

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000738

INSTRUMENT REGISTERED TO: FT WALTON BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TRUJILLO,ANTHONY,R  
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD  
ARRESTING OFFICER: TRUJILLO,ANTHONY  
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD  
LAST AGENCY INSPECTION: 03/31/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18015080A1  
DRY GAS EXPIRY DATE: 08/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 22:08  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:25

SUBJECT'S LAST NAME: NELSON

DATE/TIME OF TEST: 04/02/2016 22:53:51

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:46 |
| Air Blank          | 0.000 | 22:47 |
| Control Test       | 0.081 | 22:47 |
| Air Blank          | 0.000 | 22:48 |
| Subject Sample # 1 | 0.164 | 22:49 |
| Breath Volume      | 2.660 | 22:49 |
| Air Blank          | 0.000 | 22:49 |
| Air Blank          | 0.000 | 22:51 |
| Subject Sample # 2 | 0.156 | 22:52 |
| Breath Volume      | 2.031 | 22:52 |
| Air Blank          | 0.000 | 22:52 |
| Control Test       | 0.079 | 22:53 |
| Air Blank          | 0.000 | 22:53 |
| Diagnostics Check  | OK    | 22:53 |

---

Data Download Date/Time: 04/30/2016 12:31

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000738

INSTRUMENT REGISTERED TO: FT WALTON BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SAILOR, TRAVIS, C  
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD  
ARRESTING OFFICER: SAILOR, TRAVIS  
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD  
LAST AGENCY INSPECTION: 03/31/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18015080A1  
DRY GAS EXPIRY DATE: 08/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 19:02  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 19:45

SUBJECT'S LAST NAME: TOWNSEND

DATE/TIME OF TEST: 04/23/2016 20:24:32

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 20:18 |
| Air Blank          | 0.000 | 20:18 |
| Control Test       | 0.080 | 20:18 |
| Air Blank          | 0.000 | 20:19 |
| Subject Sample # 1 | 0.143 | 20:20 |
| Breath Volume      | 1.718 | 20:20 |
| Air Blank          | 0.000 | 20:20 |
| Air Blank          | 0.000 | 20:22 |
| Subject Sample # 2 | 0.140 | 20:23 |
| Breath Volume      | 1.667 | 20:23 |
| Air Blank          | 0.000 | 20:23 |
| Control Test       | 0.079 | 20:23 |
| Air Blank          | 0.000 | 20:24 |
| Diagnostics Check  | OK    | 20:24 |

---

Data Download Date/Time: 04/30/2016 12:31

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000738

INSTRUMENT REGISTERED TO: FT WALTON BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ESNEUL,MICHAEL,J  
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD  
ARRESTING OFFICER: ESNEUL,MICHAEL  
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD  
LAST AGENCY INSPECTION: 03/31/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18015080A1  
DRY GAS EXPIRY DATE: 08/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 19:54  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:25

SUBJECT'S LAST NAME: ONEAL-TAYLOR

DATE/TIME OF TEST: 04/29/2016 20:56:35

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 20:47 |
| Air Blank          | 0.000 | 20:47 |
| Control Test       | 0.080 | 20:48 |
| Air Blank          | 0.000 | 20:48 |
| Subject Sample # 1 | 0.177 | 20:51 |
| Breath Volume      | 1.484 | 20:51 |
| Air Blank          | 0.000 | 20:52 |
| Air Blank          | 0.000 | 20:54 |
| Subject Sample # 2 | 0.177 | 20:55 |
| Breath Volume      | 1.546 | 20:55 |
| Air Blank          | 0.000 | 20:55 |
| Control Test       | 0.078 | 20:56 |
| Air Blank          | 0.000 | 20:56 |
| Diagnostics Check  | OK    | 20:56 |

---

Data Download Date/Time: 04/30/2016 12:31

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000738

INSTRUMENT REGISTERED TO: FT WALTON BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ESNEUL,MICHAEL,J  
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD  
ARRESTING OFFICER: ESNEUL,MICHAEL  
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD  
LAST AGENCY INSPECTION: 03/31/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18015080A1  
DRY GAS EXPIRY DATE: 08/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 19:54  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:40

SUBJECT'S LAST NAME: ONEAL-TAYLOR

DATE/TIME OF TEST: 04/29/2016 21:13:53

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 21:07 |
| Air Blank          | 0.000 | 21:07 |
| Control Test       | 0.078 | 21:07 |
| Air Blank          | 0.000 | 21:08 |
| Subject Sample # 1 | 0.174 | 21:09 |
| Breath Volume      | 1.417 | 21:09 |
| Air Blank          | 0.000 | 21:09 |
| Air Blank          | 0.000 | 21:11 |
| Subject Sample # 2 | 0.176 | 21:12 |
| Breath Volume      | 2.000 | 21:12 |
| Air Blank          | 0.000 | 21:12 |
| Control Test       | 0.078 | 21:13 |
| Air Blank          | 0.000 | 21:13 |
| Diagnostics Check  | OK    | 21:13 |

---

Data Download Date/Time: 04/30/2016 12:31

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TAYLOR,KENNETH,D  
BREATH TEST OPERATOR AGENCY: Okaloosa County SO  
ARRESTING OFFICER: SEAL,JAY  
ARRESTING OFFICER AGENCY: Crestview PD  
LAST AGENCY INSPECTION: 03/29/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13115080A2  
DRY GAS EXPIRY DATE: 06/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 23:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:28

SUBJECT'S LAST NAME: FAGALIMA

DATE/TIME OF TEST: 04/03/2016 00:56:39

---

|                    |       |       |                           |
|--------------------|-------|-------|---------------------------|
| Diagnostics Check  | OK    | 00:49 |                           |
| Air Blank          | 0.000 | 00:50 |                           |
| Control Test       | 0.075 | 00:50 |                           |
| Air Blank          | 0.000 | 00:51 |                           |
| Subject Sample # 1 | 0.189 | 00:52 |                           |
| Breath Volume      | 2.390 | 00:52 |                           |
| Air Blank          | 0.000 | 00:52 |                           |
| Air Blank          | 0.000 | 00:54 |                           |
| Subject Sample # 2 | 0.174 | 00:55 |                           |
| Breath Volume      | 2.585 | 00:55 |                           |
| Air Blank          | 0.000 | 00:55 |                           |
| Control Test       | 0.072 | 00:56 | Control Outside Tolerance |
| Air Blank          | 0.000 | 00:56 |                           |

---

Data Download Date/Time: 04/28/2016 11:45

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PASCHAL,THOMAS,C  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: PASCHAL,THOMAS  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 03/29/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13115080A2  
DRY GAS EXPIRY DATE: 06/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 22:09  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: DANIEL

DATE/TIME OF TEST: 04/13/2016 23:13:09

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:06 |
| Air Blank          | 0.000 | 23:06 |
| Control Test       | 0.077 | 23:06 |
| Air Blank          | 0.000 | 23:07 |
| Subject Sample # 1 | 0.194 | 23:08 |
| Breath Volume      | 1.875 | 23:08 |
| Air Blank          | 0.000 | 23:09 |
| Air Blank          | 0.000 | 23:10 |
| Subject Sample # 2 | 0.178 | 23:11 |
| Breath Volume      | 2.000 | 23:11 |
| Air Blank          | 0.000 | 23:12 |
| Control Test       | 0.078 | 23:12 |
| Air Blank          | 0.000 | 23:13 |
| Diagnostics Check  | OK    | 23:13 |

---

Data Download Date/Time: 04/28/2016 11:45

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BOGGS,JUSTIN,E  
BREATH TEST OPERATOR AGENCY: Okaloosa County SO  
ARRESTING OFFICER: BOGGS,JUSTIN  
ARRESTING OFFICER AGENCY: Okaloosa County SO  
LAST AGENCY INSPECTION: 03/29/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13115080A2  
DRY GAS EXPIRY DATE: 06/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 22:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:14

SUBJECT'S LAST NAME: FRAZIER

DATE/TIME OF TEST: 04/15/2016 23:44:19

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:36 |
| Air Blank          | 0.000 | 23:36 |
| Control Test       | 0.078 | 23:36 |
| Air Blank          | 0.000 | 23:37 |
| Subject Sample # 1 | 0.161 | 23:39 |
| Breath Volume      | 2.656 | 23:39 |
| Air Blank          | 0.000 | 23:39 |
| Air Blank          | 0.000 | 23:41 |
| Subject Sample # 2 | 0.166 | 23:42 |
| Breath Volume      | 2.402 | 23:42 |
| Air Blank          | 0.000 | 23:43 |
| Control Test       | 0.075 | 23:43 |
| Air Blank          | 0.000 | 23:44 |
| Diagnostics Check  | OK    | 23:44 |

---

Data Download Date/Time: 04/28/2016 11:45

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RUSSAK,KEVIN,M  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: RUSSAK,KEVIN  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 03/29/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13115080A2  
DRY GAS EXPIRY DATE: 06/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 04:27  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 05:00

SUBJECT'S LAST NAME: KAYA

DATE/TIME OF TEST: 04/19/2016 05:39:59

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 05:37 |                      |
| Air Blank         | 0.000 | 05:37 |                      |
| Control Test      | 0.079 | 05:37 |                      |
| Air Blank         | 0.000 | 05:38 |                      |
| Subject Sample    | # 1   | 05:38 | Subject Test Refused |
| Breath Volume     | 0.000 | 05:38 |                      |
| Air Blank         | 0.000 | 05:39 |                      |
| Control Test      | 0.079 | 05:39 |                      |
| Air Blank         | 0.000 | 05:39 |                      |
| Diagnostics Check | OK    | 05:39 |                      |

---

Data Download Date/Time: 04/28/2016 11:45



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LIVINGSTON, RONALD, C  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: JACKSON, VICTOR  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 03/29/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13115080A2  
DRY GAS EXPIRY DATE: 06/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 02:09  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: LIVINGSTON

DATE/TIME OF TEST: 04/23/2016 02:49:24

---

|                    |       |       |                  |
|--------------------|-------|-------|------------------|
| Diagnostics Check  | OK    | 02:41 |                  |
| Air Blank          | 0.000 | 02:42 |                  |
| Control Test       | 0.078 | 02:42 |                  |
| Air Blank          | 0.000 | 02:43 |                  |
| Subject Sample # 1 | 0.236 | 02:43 |                  |
| Breath Volume      | 4.312 | 02:43 |                  |
| Air Blank          | 0.000 | 02:44 |                  |
| Air Blank          | 0.000 | 02:46 |                  |
| Subject Sample # 2 |       | 02:48 | Sequence Aborted |
| Breath Volume      | 1.394 | 02:48 |                  |
| Air Blank          | 0.000 | 02:49 |                  |

---

Data Download Date/Time: 04/28/2016 11:45

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HAND, GLENN, H  
BREATH TEST OPERATOR AGENCY: Okaloosa County SO  
ARRESTING OFFICER: HAND, GLENN  
ARRESTING OFFICER AGENCY: Okaloosa County SO  
LAST AGENCY INSPECTION: 03/29/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13115080A2  
DRY GAS EXPIRY DATE: 06/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 07:05  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 07:45

SUBJECT'S LAST NAME: KING

DATE/TIME OF TEST: 04/27/2016 08:12:20

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 08:06 |
| Air Blank          | 0.000 | 08:06 |
| Control Test       | 0.078 | 08:06 |
| Air Blank          | 0.000 | 08:07 |
| Subject Sample # 1 | 0.000 | 08:08 |
| Breath Volume      | 2.796 | 08:08 |
| Air Blank          | 0.000 | 08:08 |
| Air Blank          | 0.000 | 08:10 |
| Subject Sample # 2 | 0.000 | 08:10 |
| Breath Volume      | 1.500 | 08:10 |
| Air Blank          | 0.000 | 08:11 |
| Control Test       | 0.075 | 08:11 |
| Air Blank          | 0.000 | 08:12 |
| Diagnostics Check  | OK    | 08:12 |

---

Data Download Date/Time: 04/28/2016 11:45

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PASCHAL,THOMAS,C  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: KNOWLES,CONSTANCE  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 04/28/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13115080A2  
DRY GAS EXPIRY DATE: 06/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 14:09  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 14:55

SUBJECT'S LAST NAME: WATKINS

DATE/TIME OF TEST: 04/28/2016 15:22:41

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 15:15 |
| Air Blank          | 0.000 | 15:16 |
| Control Test       | 0.078 | 15:16 |
| Air Blank          | 0.000 | 15:17 |
| Subject Sample # 1 | 0.165 | 15:18 |
| Breath Volume      | 1.863 | 15:18 |
| Air Blank          | 0.000 | 15:18 |
| Air Blank          | 0.000 | 15:20 |
| Subject Sample # 2 | 0.176 | 15:21 |
| Breath Volume      | 1.808 | 15:21 |
| Air Blank          | 0.000 | 15:21 |
| Control Test       | 0.078 | 15:22 |
| Air Blank          | 0.000 | 15:22 |
| Diagnostics Check  | OK    | 15:22 |

---

Data Download Date/Time: 05/31/2016 10:57

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SMITH,KEVIN,R  
BREATH TEST OPERATOR AGENCY: Pensacola PD  
ARRESTING OFFICER: SMITH,KEVIN  
ARRESTING OFFICER AGENCY: Pensacola PD  
LAST AGENCY INSPECTION: 03/27/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03415080A2  
DRY GAS EXPIRY DATE: 03/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 04:06  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: RICE

DATE/TIME OF TEST: 04/01/2016 04:37:47

---

|                   |       |       |                    |
|-------------------|-------|-------|--------------------|
| Diagnostics Check | OK    | 04:26 |                    |
| Air Blank         | 0.000 | 04:26 |                    |
| Control Test      | 0.078 | 04:27 |                    |
| Air Blank         | 0.000 | 04:27 |                    |
| Subject Sample    | # 1   | 04:30 | No Sample Provided |
| Breath Volume     | 0.000 | 04:30 |                    |
| Air Blank         | 0.000 | 04:31 |                    |
| Air Blank         | 0.000 | 04:33 |                    |
| Subject Sample    | # 2   | 04:36 | No Sample Provided |
| Breath Volume     | 0.000 | 04:36 |                    |
| Air Blank         | 0.000 | 04:36 |                    |
| Control Test      | 0.080 | 04:37 |                    |
| Air Blank         | 0.000 | 04:37 |                    |
| Diagnostics Check | OK    | 04:37 |                    |

---

Data Download Date/Time: 04/13/2016 13:23

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HUDSON,JOSHUA,K  
BREATH TEST OPERATOR AGENCY: Pensacola PD  
ARRESTING OFFICER: HUDSON,JOSHUA  
ARRESTING OFFICER AGENCY: Pensacola PD  
LAST AGENCY INSPECTION: 03/27/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03415080A2  
DRY GAS EXPIRY DATE: 03/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 01:00  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:11

SUBJECT'S LAST NAME: PATE

DATE/TIME OF TEST: 04/02/2016 01:49:56

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:38 |
| Air Blank          | 0.000 | 01:38 |
| Control Test       | 0.078 | 01:39 |
| Air Blank          | 0.000 | 01:39 |
| Subject Sample # 1 | 0.231 | 01:40 |
| Breath Volume      | 1.246 | 01:40 |
| Air Blank          | 0.000 | 01:40 |
| Air Blank          | 0.000 | 01:42 |
| Subject Sample # 2 | 0.252 | 01:43 |
| Breath Volume      | 1.171 | 01:43 |
| Air Blank          | 0.000 | 01:44 |
| Air Blank          | 0.000 | 01:46 |
| Subject Sample # 3 | 0.223 | 01:48 |
| Breath Volume      | 2.000 | 01:48 |
| Air Blank          | 0.000 | 01:49 |
| Control Test       | 0.077 | 01:49 |
| Air Blank          | 0.000 | 01:49 |
| Diagnostics Check  | OK    | 01:49 |

---

Data Download Date/Time: 04/13/2016 13:23

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WRIGHT,L,  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: WRIGHT,L  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 03/27/2016

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 03415080A2  
DRY GAS EXPIRY DATE: 03/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 19:42  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 19:55

SUBJECT'S LAST NAME: SLAUGHTER

DATE/TIME OF TEST: 04/02/2016 20:31:17

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 20:21 |
| Air Blank          | 0.000 | 20:22 |
| Control Test       | 0.078 | 20:22 |
| Air Blank          | 0.000 | 20:23 |
| Subject Sample # 1 | 0.237 | 20:23 |
| Breath Volume      | 2.531 | 20:23 |
| Air Blank          | 0.000 | 20:24 |
| Air Blank          | 0.000 | 20:26 |
| Subject Sample # 2 | 0.213 | 20:26 |
| Breath Volume      | 1.593 | 20:26 |
| Air Blank          | 0.000 | 20:27 |
| Air Blank          | 0.000 | 20:29 |
| Subject Sample # 3 | 0.222 | 20:29 |
| Breath Volume      | 2.445 | 20:29 |
| Air Blank          | 0.000 | 20:30 |
| Control Test       | 0.077 | 20:30 |
| Air Blank          | 0.000 | 20:31 |
| Diagnostics Check  | OK    | 20:31 |

---

Data Download Date/Time: 04/13/2016 13:23

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SHORETTE,BRIAN,E  
BREATH TEST OPERATOR AGENCY: Escambia County SO  
ARRESTING OFFICER: SHORETTE,BRIAN  
ARRESTING OFFICER AGENCY: Escambia County SO  
LAST AGENCY INSPECTION: 03/27/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03415080A2  
DRY GAS EXPIRY DATE: 03/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 00:26  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:43

SUBJECT'S LAST NAME: CARTER

DATE/TIME OF TEST: 04/03/2016 01:09:42

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:03 |
| Air Blank          | 0.000 | 01:04 |
| Control Test       | 0.079 | 01:04 |
| Air Blank          | 0.000 | 01:04 |
| Subject Sample # 1 | 0.177 | 01:05 |
| Breath Volume      | 2.652 | 01:05 |
| Air Blank          | 0.000 | 01:06 |
| Air Blank          | 0.000 | 01:07 |
| Subject Sample # 2 | 0.179 | 01:08 |
| Breath Volume      | 2.093 | 01:08 |
| Air Blank          | 0.000 | 01:08 |
| Control Test       | 0.078 | 01:09 |
| Air Blank          | 0.000 | 01:09 |
| Diagnostics Check  | OK    | 01:09 |

---

Data Download Date/Time: 04/13/2016 13:23

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HUDSON,JOSHUA,K  
BREATH TEST OPERATOR AGENCY: Pensacola PD  
ARRESTING OFFICER: HUDSON,JOSHUA  
ARRESTING OFFICER AGENCY: Pensacola PD  
LAST AGENCY INSPECTION: 03/27/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03415080A2  
DRY GAS EXPIRY DATE: 03/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 02:51  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: BARBER

DATE/TIME OF TEST: 04/03/2016 03:35:52

---

|                   |       |       |                    |
|-------------------|-------|-------|--------------------|
| Diagnostics Check | OK    | 03:24 |                    |
| Air Blank         | 0.000 | 03:24 |                    |
| Control Test      | 0.078 | 03:25 |                    |
| Air Blank         | 0.000 | 03:25 |                    |
| Subject Sample    | # 1   | 03:28 | No Sample Provided |
| Breath Volume     | 0.000 | 03:28 |                    |
| Air Blank         | 0.000 | 03:29 |                    |
| Air Blank         | 0.000 | 03:31 |                    |
| Subject Sample    | # 2   | 03:34 | No Sample Provided |
| Breath Volume     | 0.000 | 03:34 |                    |
| Air Blank         | 0.000 | 03:34 |                    |
| Control Test      | 0.078 | 03:35 |                    |
| Air Blank         | 0.000 | 03:35 |                    |
| Diagnostics Check | OK    | 03:35 |                    |

---

Data Download Date/Time: 04/13/2016 13:23



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SHELLEY,SAMUEL,L  
BREATH TEST OPERATOR AGENCY: Escambia County SO  
ARRESTING OFFICER: HENLEY,WILLIAM  
ARRESTING OFFICER AGENCY: Escambia County SO  
LAST AGENCY INSPECTION: 03/27/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03415080A2  
DRY GAS EXPIRY DATE: 03/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 21:36  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:08

SUBJECT'S LAST NAME: COATES

DATE/TIME OF TEST: 04/04/2016 22:35:19

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:28 |
| Air Blank          | 0.000 | 22:28 |
| Control Test       | 0.079 | 22:29 |
| Air Blank          | 0.000 | 22:29 |
| Subject Sample # 1 | 0.025 | 22:30 |
| Breath Volume      | 1.203 | 22:30 |
| Air Blank          | 0.000 | 22:31 |
| Air Blank          | 0.000 | 22:33 |
| Subject Sample # 2 | 0.026 | 22:33 |
| Breath Volume      | 1.472 | 22:33 |
| Air Blank          | 0.000 | 22:34 |
| Control Test       | 0.078 | 22:34 |
| Air Blank          | 0.000 | 22:35 |
| Diagnostics Check  | OK    | 22:35 |

---

Data Download Date/Time: 04/13/2016 13:23

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HUDSON,JOSHUA,K  
BREATH TEST OPERATOR AGENCY: Pensacola PD  
ARRESTING OFFICER: HUDSON,JOSHUA  
ARRESTING OFFICER AGENCY: Pensacola PD  
LAST AGENCY INSPECTION: 03/27/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03415080A2  
DRY GAS EXPIRY DATE: 03/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 01:56  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: BURD

DATE/TIME OF TEST: 04/05/2016 02:51:03

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:43 |
| Air Blank          | 0.000 | 02:43 |
| Control Test       | 0.078 | 02:44 |
| Air Blank          | 0.000 | 02:44 |
| Subject Sample # 1 | 0.181 | 02:46 |
| Breath Volume      | 1.890 | 02:46 |
| Air Blank          | 0.000 | 02:47 |
| Air Blank          | 0.000 | 02:48 |
| Subject Sample # 2 | 0.173 | 02:49 |
| Breath Volume      | 1.613 | 02:49 |
| Air Blank          | 0.000 | 02:50 |
| Control Test       | 0.077 | 02:50 |
| Air Blank          | 0.000 | 02:50 |
| Diagnostics Check  | OK    | 02:51 |

---

Data Download Date/Time: 04/13/2016 13:23

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SHORETTE,BRIAN,E  
BREATH TEST OPERATOR AGENCY: Escambia County SO  
ARRESTING OFFICER: HOLCOMBE,JOHN  
ARRESTING OFFICER AGENCY: Escambia County SO  
LAST AGENCY INSPECTION: 03/27/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03415080A2  
DRY GAS EXPIRY DATE: 03/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 18:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:27

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 04/05/2016 18:54:05

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 18:47 |
| Air Blank          | 0.000 | 18:48 |
| Control Test       | 0.078 | 18:48 |
| Air Blank          | 0.000 | 18:49 |
| Subject Sample # 1 | 0.098 | 18:49 |
| Breath Volume      | 1.953 | 18:49 |
| Air Blank          | 0.000 | 18:50 |
| Air Blank          | 0.000 | 18:52 |
| Subject Sample # 2 | 0.082 | 18:52 |
| Breath Volume      | 1.425 | 18:52 |
| Air Blank          | 0.000 | 18:53 |
| Control Test       | 0.077 | 18:53 |
| Air Blank          | 0.000 | 18:54 |
| Diagnostics Check  | OK    | 18:54 |

---

Data Download Date/Time: 04/13/2016 13:23

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HUDSON,JOSHUA,K  
BREATH TEST OPERATOR AGENCY: Pensacola PD  
ARRESTING OFFICER: HUDSON,JOSHUA  
ARRESTING OFFICER AGENCY: Pensacola PD  
LAST AGENCY INSPECTION: 03/27/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03415080A2  
DRY GAS EXPIRY DATE: 03/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 00:39  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 04/08/2016 01:23:10

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:16 |
| Air Blank          | 0.000 | 01:16 |
| Control Test       | 0.079 | 01:17 |
| Air Blank          | 0.000 | 01:17 |
| Subject Sample # 1 | 0.113 | 01:18 |
| Breath Volume      | 2.425 | 01:18 |
| Air Blank          | 0.000 | 01:19 |
| Air Blank          | 0.000 | 01:21 |
| Subject Sample # 2 | 0.110 | 01:21 |
| Breath Volume      | 2.355 | 01:21 |
| Air Blank          | 0.000 | 01:22 |
| Control Test       | 0.080 | 01:22 |
| Air Blank          | 0.000 | 01:23 |
| Diagnostics Check  | OK    | 01:23 |

---

Data Download Date/Time: 04/13/2016 13:23

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DANIELS,JAMES,A  
BREATH TEST OPERATOR AGENCY: Pensacola PD  
ARRESTING OFFICER: DANIELS,JAMES  
ARRESTING OFFICER AGENCY: Pensacola PD  
LAST AGENCY INSPECTION: 03/27/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03415080A2  
DRY GAS EXPIRY DATE: 03/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 02:24  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: ZAVADA

DATE/TIME OF TEST: 04/08/2016 02:57:44

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:50 |
| Air Blank          | 0.000 | 02:50 |
| Control Test       | 0.077 | 02:50 |
| Air Blank          | 0.000 | 02:51 |
| Subject Sample # 1 | 0.221 | 02:52 |
| Breath Volume      | 2.304 | 02:52 |
| Air Blank          | 0.000 | 02:53 |
| Air Blank          | 0.000 | 02:55 |
| Subject Sample # 2 | 0.210 | 02:56 |
| Breath Volume      | 1.679 | 02:56 |
| Air Blank          | 0.000 | 02:56 |
| Control Test       | 0.078 | 02:57 |
| Air Blank          | 0.000 | 02:57 |
| Diagnostics Check  | OK    | 02:57 |

---

Data Download Date/Time: 04/13/2016 13:23

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DANIELS,JAMES,A  
BREATH TEST OPERATOR AGENCY: Pensacola PD  
ARRESTING OFFICER: DANIELS,JAMES  
ARRESTING OFFICER AGENCY: Pensacola PD  
LAST AGENCY INSPECTION: 03/27/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03415080A2  
DRY GAS EXPIRY DATE: 03/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 23:47  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: CURRAN

DATE/TIME OF TEST: 04/09/2016 00:24:13

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:16 |
| Air Blank          | 0.000 | 00:16 |
| Control Test       | 0.079 | 00:17 |
| Air Blank          | 0.000 | 00:17 |
| Subject Sample # 1 | 0.129 | 00:19 |
| Breath Volume      | 1.171 | 00:19 |
| Air Blank          | 0.000 | 00:20 |
| Air Blank          | 0.000 | 00:21 |
| Subject Sample # 2 | 0.121 | 00:22 |
| Breath Volume      | 2.468 | 00:22 |
| Air Blank          | 0.000 | 00:23 |
| Control Test       | 0.079 | 00:23 |
| Air Blank          | 0.000 | 00:24 |
| Diagnostics Check  | OK    | 00:24 |

---

Data Download Date/Time: 04/13/2016 13:23

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HUDSON,JOSHUA,K  
BREATH TEST OPERATOR AGENCY: Pensacola PD  
ARRESTING OFFICER: HUDSON,JOSHUA  
ARRESTING OFFICER AGENCY: Pensacola PD  
LAST AGENCY INSPECTION: 03/27/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03415080A2  
DRY GAS EXPIRY DATE: 03/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 02:32  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:34

SUBJECT'S LAST NAME: DAVILA PATRELIS

DATE/TIME OF TEST: 04/09/2016 03:09:05

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 03:03 |                      |
| Air Blank         | 0.000 | 03:04 |                      |
| Control Test      | 0.078 | 03:04 |                      |
| Air Blank         | 0.000 | 03:05 |                      |
| Subject Sample    | # 1   | 03:07 | Subject Test Refused |
| Breath Volume     | 0.000 | 03:07 |                      |
| Air Blank         | 0.000 | 03:08 |                      |
| Control Test      | 0.079 | 03:08 |                      |
| Air Blank         | 0.000 | 03:08 |                      |
| Diagnostics Check | OK    | 03:09 |                      |

---

Data Download Date/Time: 04/13/2016 13:23

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PILLOW, JEREMY, T  
BREATH TEST OPERATOR AGENCY: Pensacola PD  
ARRESTING OFFICER: FOX, ALEX  
ARRESTING OFFICER AGENCY: Pensacola PD  
LAST AGENCY INSPECTION: 03/27/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03415080A2  
DRY GAS EXPIRY DATE: 03/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 03:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:54

SUBJECT'S LAST NAME: PICADO VARGAS

DATE/TIME OF TEST: 04/09/2016 04:21:12

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:14 |
| Air Blank          | 0.000 | 04:14 |
| Control Test       | 0.078 | 04:15 |
| Air Blank          | 0.000 | 04:15 |
| Subject Sample # 1 | 0.104 | 04:16 |
| Breath Volume      | 1.964 | 04:16 |
| Air Blank          | 0.000 | 04:17 |
| Air Blank          | 0.000 | 04:19 |
| Subject Sample # 2 | 0.093 | 04:19 |
| Breath Volume      | 1.417 | 04:19 |
| Air Blank          | 0.000 | 04:20 |
| Control Test       | 0.078 | 04:20 |
| Air Blank          | 0.000 | 04:21 |
| Diagnostics Check  | OK    | 04:21 |

---

Data Download Date/Time: 04/13/2016 13:23



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HUDSON,JOSHUA,K  
BREATH TEST OPERATOR AGENCY: Pensacola PD  
ARRESTING OFFICER: HUDSON,JOSHUA  
ARRESTING OFFICER AGENCY: Pensacola PD  
LAST AGENCY INSPECTION: 03/27/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03415080A2  
DRY GAS EXPIRY DATE: 03/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 23:29  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:48

SUBJECT'S LAST NAME: LYNN

DATE/TIME OF TEST: 04/10/2016 00:27:04

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:20 |
| Air Blank          | 0.000 | 00:21 |
| Control Test       | 0.079 | 00:21 |
| Air Blank          | 0.000 | 00:22 |
| Subject Sample # 1 | 0.000 | 00:22 |
| Breath Volume      | 2.429 | 00:22 |
| Air Blank          | 0.000 | 00:23 |
| Air Blank          | 0.000 | 00:25 |
| Subject Sample # 2 | 0.000 | 00:25 |
| Breath Volume      | 3.152 | 00:25 |
| Air Blank          | 0.000 | 00:26 |
| Control Test       | 0.079 | 00:26 |
| Air Blank          | 0.000 | 00:26 |
| Diagnostics Check  | OK    | 00:27 |

---

Data Download Date/Time: 04/13/2016 13:23

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SHELLEY,SAMUEL,L  
BREATH TEST OPERATOR AGENCY: Escambia County SO  
ARRESTING OFFICER: SHELLEY,SAMUEL  
ARRESTING OFFICER AGENCY: Escambia County SO  
LAST AGENCY INSPECTION: 03/27/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03415080A2  
DRY GAS EXPIRY DATE: 03/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 01:33  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: ROBINSON

DATE/TIME OF TEST: 04/10/2016 02:27:13

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:21 |
| Air Blank          | 0.000 | 02:21 |
| Control Test       | 0.078 | 02:21 |
| Air Blank          | 0.000 | 02:22 |
| Subject Sample # 1 | 0.161 | 02:22 |
| Breath Volume      | 1.496 | 02:22 |
| Air Blank          | 0.000 | 02:23 |
| Air Blank          | 0.000 | 02:25 |
| Subject Sample # 2 | 0.171 | 02:25 |
| Breath Volume      | 1.644 | 02:25 |
| Air Blank          | 0.000 | 02:26 |
| Control Test       | 0.077 | 02:26 |
| Air Blank          | 0.000 | 02:27 |
| Diagnostics Check  | OK    | 02:27 |

---

Data Download Date/Time: 04/13/2016 13:23

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HUDSON,JOSHUA,K  
BREATH TEST OPERATOR AGENCY: Pensacola PD  
ARRESTING OFFICER: RESTIFO,CHARLES  
ARRESTING OFFICER AGENCY: Pensacola PD  
LAST AGENCY INSPECTION: 03/27/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03415080A2  
DRY GAS EXPIRY DATE: 03/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 02:21  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:41

SUBJECT'S LAST NAME: OSBORNE

DATE/TIME OF TEST: 04/10/2016 03:13:17

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:06 |
| Air Blank          | 0.000 | 03:06 |
| Control Test       | 0.078 | 03:06 |
| Air Blank          | 0.000 | 03:07 |
| Subject Sample # 1 | 0.167 | 03:08 |
| Breath Volume      | 1.792 | 03:08 |
| Air Blank          | 0.000 | 03:08 |
| Air Blank          | 0.000 | 03:10 |
| Subject Sample # 2 | 0.160 | 03:11 |
| Breath Volume      | 2.878 | 03:11 |
| Air Blank          | 0.000 | 03:12 |
| Control Test       | 0.079 | 03:12 |
| Air Blank          | 0.000 | 03:13 |
| Diagnostics Check  | OK    | 03:13 |

---

Data Download Date/Time: 04/13/2016 13:23

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HUDSON,JOSHUA,K  
BREATH TEST OPERATOR AGENCY: Pensacola PD  
ARRESTING OFFICER: BYRD,BRANDON  
ARRESTING OFFICER AGENCY: Pensacola PD  
LAST AGENCY INSPECTION: 03/27/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03415080A2  
DRY GAS EXPIRY DATE: 03/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 02:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:57

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 04/10/2016 03:45:02

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:37 |
| Air Blank          | 0.000 | 03:37 |
| Control Test       | 0.078 | 03:38 |
| Air Blank          | 0.000 | 03:38 |
| Subject Sample # 1 | 0.128 | 03:40 |
| Breath Volume      | 3.523 | 03:40 |
| Air Blank          | 0.000 | 03:40 |
| Air Blank          | 0.000 | 03:42 |
| Subject Sample # 2 | 0.129 | 03:43 |
| Breath Volume      | 2.792 | 03:43 |
| Air Blank          | 0.000 | 03:44 |
| Control Test       | 0.078 | 03:44 |
| Air Blank          | 0.000 | 03:44 |
| Diagnostics Check  | OK    | 03:45 |

---

Data Download Date/Time: 04/13/2016 13:23

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PILLOW, JEREMY, T  
BREATH TEST OPERATOR AGENCY: Pensacola PD  
ARRESTING OFFICER: PILLOW, JEREMY  
ARRESTING OFFICER AGENCY: Pensacola PD  
LAST AGENCY INSPECTION: 03/27/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03415080A2  
DRY GAS EXPIRY DATE: 03/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 20:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:24

SUBJECT'S LAST NAME: ENOS

DATE/TIME OF TEST: 04/10/2016 20:52:02

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 20:45 |
| Air Blank          | 0.000 | 20:45 |
| Control Test       | 0.078 | 20:46 |
| Air Blank          | 0.000 | 20:46 |
| Subject Sample # 1 | 0.203 | 20:47 |
| Breath Volume      | 1.500 | 20:47 |
| Air Blank          | 0.000 | 20:48 |
| Air Blank          | 0.000 | 20:49 |
| Subject Sample # 2 | 0.202 | 20:50 |
| Breath Volume      | 1.167 | 20:50 |
| Air Blank          | 0.000 | 20:51 |
| Control Test       | 0.077 | 20:51 |
| Air Blank          | 0.000 | 20:51 |
| Diagnostics Check  | OK    | 20:52 |

---

Data Download Date/Time: 04/13/2016 13:23

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: REGISTER, MICHAEL, A  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: REGISTER, MICHAEL  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 03/27/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03415080A2  
DRY GAS EXPIRY DATE: 03/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 21:34  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: VINES

DATE/TIME OF TEST: 04/10/2016 22:33:58

---

|                   |           |       |                      |
|-------------------|-----------|-------|----------------------|
| Diagnostics Check | OK        | 22:24 |                      |
| Air Blank         | 0.000     | 22:25 |                      |
| Control Test      | 0.079     | 22:25 |                      |
| Air Blank         | 0.000     | 22:26 |                      |
| Subject Sample    | # 1 0.266 | 22:29 | Volume Not Met       |
| Breath Volume     | 0.925     | 22:29 |                      |
| Air Blank         | 0.000     | 22:30 |                      |
| Air Blank         | 0.000     | 22:31 |                      |
| Subject Sample    | # 2       | 22:32 | Subject Test Refused |
| Breath Volume     | 0.687     | 22:32 |                      |
| Air Blank         | 0.000     | 22:33 |                      |
| Control Test      | 0.076     | 22:33 |                      |
| Air Blank         | 0.000     | 22:33 |                      |
| Diagnostics Check | OK        | 22:33 |                      |

---

Data Download Date/Time: 04/13/2016 13:23

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HUDSON,JOSHUA,K  
BREATH TEST OPERATOR AGENCY: Pensacola PD  
ARRESTING OFFICER: HUDSON,JOSHUA  
ARRESTING OFFICER AGENCY: Pensacola PD  
LAST AGENCY INSPECTION: 03/27/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03415080A2  
DRY GAS EXPIRY DATE: 03/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 03:29  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:38

SUBJECT'S LAST NAME: BAUMBACH

DATE/TIME OF TEST: 04/13/2016 04:14:10

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:06 |
| Air Blank          | 0.000 | 04:07 |
| Control Test       | 0.079 | 04:07 |
| Air Blank          | 0.000 | 04:08 |
| Subject Sample # 1 | 0.179 | 04:09 |
| Breath Volume      | 1.906 | 04:09 |
| Air Blank          | 0.000 | 04:10 |
| Air Blank          | 0.000 | 04:11 |
| Subject Sample # 2 | 0.189 | 04:12 |
| Breath Volume      | 1.375 | 04:12 |
| Air Blank          | 0.000 | 04:13 |
| Control Test       | 0.079 | 04:13 |
| Air Blank          | 0.000 | 04:14 |
| Diagnostics Check  | OK    | 04:14 |

---

Data Download Date/Time: 04/13/2016 13:23

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PILLOW, JEREMY, T  
BREATH TEST OPERATOR AGENCY: Pensacola PD  
ARRESTING OFFICER: SIEMEN, DAN  
ARRESTING OFFICER AGENCY: Pensacola PD  
LAST AGENCY INSPECTION: 04/13/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03415080A2  
DRY GAS EXPIRY DATE: 03/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 17:50  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 18:06

SUBJECT'S LAST NAME: HUNTER

DATE/TIME OF TEST: 04/13/2016 18:39:18

---

|                   |       |       |                    |
|-------------------|-------|-------|--------------------|
| Diagnostics Check | OK    | 18:27 |                    |
| Air Blank         | 0.000 | 18:28 |                    |
| Control Test      | 0.079 | 18:28 |                    |
| Air Blank         | 0.000 | 18:29 |                    |
| Subject Sample    | # 1   | 18:32 | No Sample Provided |
| Breath Volume     | 0.000 | 18:32 |                    |
| Air Blank         | 0.000 | 18:32 |                    |
| Air Blank         | 0.000 | 18:34 |                    |
| Subject Sample    | # 2   | 18:37 | No Sample Provided |
| Breath Volume     | 0.000 | 18:37 |                    |
| Air Blank         | 0.000 | 18:38 |                    |
| Control Test      | 0.079 | 18:38 |                    |
| Air Blank         | 0.000 | 18:39 |                    |
| Diagnostics Check | OK    | 18:39 |                    |

---

Data Download Date/Time: 05/21/2016 04:49



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PILLOW, JEREMY, T  
BREATH TEST OPERATOR AGENCY: Pensacola PD  
ARRESTING OFFICER: SIEMEN, DAN  
ARRESTING OFFICER AGENCY: Pensacola PD  
LAST AGENCY INSPECTION: 04/13/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03415080A2  
DRY GAS EXPIRY DATE: 03/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 17:40  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 18:08

SUBJECT'S LAST NAME: HUNTER

DATE/TIME OF TEST: 04/13/2016 18:54:51

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 18:41 |
| Air Blank          | 0.000 | 18:42 |
| Control Test       | 0.079 | 18:42 |
| Air Blank          | 0.000 | 18:43 |
| Subject Sample # 1 | 0.178 | 18:45 |
| Breath Volume      | 1.296 | 18:45 |
| Air Blank          | 0.000 | 18:46 |
| Air Blank          | 0.000 | 18:48 |
| Subject Sample # 2 | 0.200 | 18:49 |
| Breath Volume      | 1.234 | 18:49 |
| Air Blank          | 0.000 | 18:50 |
| Air Blank          | 0.000 | 18:52 |
| Subject Sample # 3 | 0.197 | 18:53 |
| Breath Volume      | 1.449 | 18:53 |
| Air Blank          | 0.000 | 18:53 |
| Control Test       | 0.076 | 18:54 |
| Air Blank          | 0.000 | 18:54 |
| Diagnostics Check  | OK    | 18:54 |

---

Data Download Date/Time: 05/21/2016 04:49

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SHELLEY,SAMUEL,L  
BREATH TEST OPERATOR AGENCY: Escambia County SO  
ARRESTING OFFICER: SHELLEY,SAMUEL  
ARRESTING OFFICER AGENCY: Escambia County SO  
LAST AGENCY INSPECTION: 04/13/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03415080A2  
DRY GAS EXPIRY DATE: 03/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 23:31  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: KASHUNG

DATE/TIME OF TEST: 04/14/2016 00:29:24

---

|                   |           |       |                |
|-------------------|-----------|-------|----------------|
| Diagnostics Check | OK        | 00:17 |                |
| Air Blank         | 0.000     | 00:18 |                |
| Control Test      | 0.078     | 00:18 |                |
| Air Blank         | 0.000     | 00:18 |                |
| Subject Sample    | # 1 0.199 | 00:22 | Volume Not Met |
| Breath Volume     | 0.308     | 00:22 |                |
| Air Blank         | 0.000     | 00:22 |                |
| Air Blank         | 0.000     | 00:24 |                |
| Subject Sample    | # 2 0.232 | 00:27 | Volume Not Met |
| Breath Volume     | 0.601     | 00:27 |                |
| Air Blank         | 0.000     | 00:28 |                |
| Control Test      | 0.078     | 00:28 |                |
| Air Blank         | 0.000     | 00:29 |                |
| Diagnostics Check | OK        | 00:29 |                |

---

Data Download Date/Time: 05/21/2016 04:49

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SHELLEY,SAMUEL,L  
BREATH TEST OPERATOR AGENCY: Escambia County SO  
ARRESTING OFFICER: SHELLEY,SAMUEL  
ARRESTING OFFICER AGENCY: Escambia County SO  
LAST AGENCY INSPECTION: 04/13/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03415080A2  
DRY GAS EXPIRY DATE: 03/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 23:31  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: KASHUNG

DATE/TIME OF TEST: 04/14/2016 00:40:53

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:31 |
| Air Blank          | 0.000 | 00:32 |
| Control Test       | 0.077 | 00:32 |
| Air Blank          | 0.000 | 00:33 |
| Subject Sample # 1 | 0.231 | 00:34 |
| Breath Volume      | 1.515 | 00:34 |
| Air Blank          | 0.000 | 00:34 |
| Air Blank          | 0.000 | 00:36 |
| Subject Sample # 2 | 0.223 | 00:39 |
| Breath Volume      | 1.199 | 00:39 |
| Air Blank          | 0.000 | 00:39 |
| Control Test       | 0.077 | 00:40 |
| Air Blank          | 0.000 | 00:40 |
| Diagnostics Check  | OK    | 00:40 |

---

Data Download Date/Time: 05/21/2016 04:49

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SHELLEY,SAMUEL,L  
BREATH TEST OPERATOR AGENCY: Escambia County SO  
ARRESTING OFFICER: SHELLEY,SAMUEL  
ARRESTING OFFICER AGENCY: Escambia County SO  
LAST AGENCY INSPECTION: 04/13/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03415080A2  
DRY GAS EXPIRY DATE: 03/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 22:47  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 04/14/2016 23:47:02

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:40 |
| Air Blank          | 0.000 | 23:41 |
| Control Test       | 0.079 | 23:41 |
| Air Blank          | 0.000 | 23:42 |
| Subject Sample # 1 | 0.000 | 23:42 |
| Breath Volume      | 1.500 | 23:42 |
| Air Blank          | 0.000 | 23:43 |
| Air Blank          | 0.000 | 23:45 |
| Subject Sample # 2 | 0.000 | 23:45 |
| Breath Volume      | 1.394 | 23:45 |
| Air Blank          | 0.000 | 23:46 |
| Control Test       | 0.079 | 23:46 |
| Air Blank          | 0.000 | 23:46 |
| Diagnostics Check  | OK    | 23:47 |

---

Data Download Date/Time: 05/21/2016 04:49

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HUDSON,JOSHUA,K  
BREATH TEST OPERATOR AGENCY: Pensacola PD  
ARRESTING OFFICER: HUDSON,JOSHUA  
ARRESTING OFFICER AGENCY: Pensacola PD  
LAST AGENCY INSPECTION: 04/13/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03415080A2  
DRY GAS EXPIRY DATE: 03/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 01:07  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:09

SUBJECT'S LAST NAME: CASTAGNA

DATE/TIME OF TEST: 04/15/2016 01:44:16

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:37 |
| Air Blank          | 0.000 | 01:37 |
| Control Test       | 0.077 | 01:37 |
| Air Blank          | 0.000 | 01:38 |
| Subject Sample # 1 | 0.187 | 01:39 |
| Breath Volume      | 2.785 | 01:39 |
| Air Blank          | 0.000 | 01:39 |
| Air Blank          | 0.000 | 01:41 |
| Subject Sample # 2 | 0.198 | 01:42 |
| Breath Volume      | 2.316 | 01:42 |
| Air Blank          | 0.000 | 01:43 |
| Control Test       | 0.077 | 01:43 |
| Air Blank          | 0.000 | 01:44 |
| Diagnostics Check  | OK    | 01:44 |

---

Data Download Date/Time: 05/21/2016 04:49

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SHORETTE,BRIAN,E  
BREATH TEST OPERATOR AGENCY: Escambia County SO  
ARRESTING OFFICER: SHORETTE,BRIAN  
ARRESTING OFFICER AGENCY: Escambia County SO  
LAST AGENCY INSPECTION: 04/13/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03415080A2  
DRY GAS EXPIRY DATE: 03/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 23:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:49

SUBJECT'S LAST NAME: ORMAND

DATE/TIME OF TEST: 04/16/2016 00:15:31

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:09 |
| Air Blank          | 0.000 | 00:09 |
| Control Test       | 0.080 | 00:10 |
| Air Blank          | 0.000 | 00:10 |
| Subject Sample # 1 | 0.162 | 00:11 |
| Breath Volume      | 2.554 | 00:11 |
| Air Blank          | 0.000 | 00:11 |
| Air Blank          | 0.000 | 00:13 |
| Subject Sample # 2 | 0.158 | 00:13 |
| Breath Volume      | 2.687 | 00:13 |
| Air Blank          | 0.000 | 00:14 |
| Control Test       | 0.078 | 00:14 |
| Air Blank          | 0.000 | 00:15 |
| Diagnostics Check  | OK    | 00:15 |

---

Data Download Date/Time: 05/21/2016 04:49

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BRUCE, HEATHER, N  
BREATH TEST OPERATOR AGENCY: Escambia County SO  
ARRESTING OFFICER: BRUCE, HEATHER  
ARRESTING OFFICER AGENCY: Escambia County SO  
LAST AGENCY INSPECTION: 04/13/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03415080A2  
DRY GAS EXPIRY DATE: 03/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 18:27  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:02

SUBJECT'S LAST NAME: HENRY

DATE/TIME OF TEST: 04/17/2016 19:29:25

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 19:26 |                      |
| Air Blank         | 0.000 | 19:27 |                      |
| Control Test      | 0.077 | 19:27 |                      |
| Air Blank         | 0.000 | 19:27 |                      |
| Subject Sample    | # 1   | 19:28 | Subject Test Refused |
| Breath Volume     | 0.000 | 19:28 |                      |
| Air Blank         | 0.000 | 19:28 |                      |
| Control Test      | 0.077 | 19:28 |                      |
| Air Blank         | 0.000 | 19:29 |                      |
| Diagnostics Check | OK    | 19:29 |                      |

---

Data Download Date/Time: 05/21/2016 04:49

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DANIELS,JAMES,A  
BREATH TEST OPERATOR AGENCY: Pensacola PD  
ARRESTING OFFICER: BARBOSA,SARAH  
ARRESTING OFFICER AGENCY: Pensacola PD  
LAST AGENCY INSPECTION: 04/13/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03415080A2  
DRY GAS EXPIRY DATE: 03/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 22:46  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: CUNNINGHAM

DATE/TIME OF TEST: 04/22/2016 23:30:04

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 23:20 |                      |
| Air Blank         | 0.000 | 23:20 |                      |
| Control Test      | 0.079 | 23:21 |                      |
| Air Blank         | 0.000 | 23:21 |                      |
| Subject Sample    | # 1   | 23:24 | No Sample Provided   |
| Breath Volume     | 0.000 | 23:24 |                      |
| Air Blank         | 0.000 | 23:25 |                      |
| Air Blank         | 0.000 | 23:27 |                      |
| Subject Sample    | # 2   | 23:28 | Subject Test Refused |
| Breath Volume     | 0.000 | 23:28 |                      |
| Air Blank         | 0.000 | 23:29 |                      |
| Control Test      | 0.078 | 23:29 |                      |
| Air Blank         | 0.000 | 23:29 |                      |
| Diagnostics Check | OK    | 23:30 |                      |

---

Data Download Date/Time: 05/21/2016 04:49



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: STEVENS, MITCHELL, M  
BREATH TEST OPERATOR AGENCY: Escambia County SO  
ARRESTING OFFICER: STEVENS, MITCHELL  
ARRESTING OFFICER AGENCY: Escambia County SO  
LAST AGENCY INSPECTION: 04/13/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03415080A2  
DRY GAS EXPIRY DATE: 03/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 01:16  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:38

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 04/23/2016 02:05:15

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:58 |
| Air Blank          | 0.000 | 01:59 |
| Control Test       | 0.079 | 01:59 |
| Air Blank          | 0.000 | 01:59 |
| Subject Sample # 1 | 0.000 | 02:00 |
| Breath Volume      | 2.605 | 02:00 |
| Air Blank          | 0.000 | 02:01 |
| Air Blank          | 0.000 | 02:03 |
| Subject Sample # 2 | 0.000 | 02:03 |
| Breath Volume      | 2.800 | 02:03 |
| Air Blank          | 0.000 | 02:04 |
| Control Test       | 0.079 | 02:04 |
| Air Blank          | 0.000 | 02:05 |
| Diagnostics Check  | OK    | 02:05 |

---

Data Download Date/Time: 05/21/2016 04:49

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HUDSON,JOSHUA,K  
BREATH TEST OPERATOR AGENCY: Pensacola PD  
ARRESTING OFFICER: HUDSON,JOSHUA  
ARRESTING OFFICER AGENCY: Pensacola PD  
LAST AGENCY INSPECTION: 04/13/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03415080A2  
DRY GAS EXPIRY DATE: 03/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 21:08  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: MORRIS

DATE/TIME OF TEST: 04/25/2016 22:00:20

---

|                   |       |       |                    |
|-------------------|-------|-------|--------------------|
| Diagnostics Check | OK    | 21:48 |                    |
| Air Blank         | 0.000 | 21:49 |                    |
| Control Test      | 0.078 | 21:49 |                    |
| Air Blank         | 0.000 | 21:50 |                    |
| Subject Sample    | # 1   | 21:53 | No Sample Provided |
| Breath Volume     | 0.000 | 21:53 |                    |
| Air Blank         | 0.000 | 21:53 |                    |
| Air Blank         | 0.000 | 21:55 |                    |
| Subject Sample    | # 2   | 21:58 | No Sample Provided |
| Breath Volume     | 0.000 | 21:58 |                    |
| Air Blank         | 0.000 | 21:59 |                    |
| Control Test      | 0.078 | 21:59 |                    |
| Air Blank         | 0.000 | 22:00 |                    |
| Diagnostics Check | OK    | 22:00 |                    |

---

Data Download Date/Time: 05/21/2016 04:49

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FLEMING,ROBERT,C  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: VO,SEBASTIAN  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 04/13/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03415080A2  
DRY GAS EXPIRY DATE: 03/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 21:12  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:41

SUBJECT'S LAST NAME: FLEMING

DATE/TIME OF TEST: 04/25/2016 22:13:39

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:07 |
| Air Blank          | 0.000 | 22:08 |
| Control Test       | 0.078 | 22:08 |
| Air Blank          | 0.000 | 22:08 |
| Subject Sample # 1 | 0.164 | 22:09 |
| Breath Volume      | 1.703 | 22:09 |
| Air Blank          | 0.000 | 22:09 |
| Air Blank          | 0.000 | 22:11 |
| Subject Sample # 2 | 0.160 | 22:12 |
| Breath Volume      | 1.773 | 22:12 |
| Air Blank          | 0.000 | 22:12 |
| Control Test       | 0.078 | 22:13 |
| Air Blank          | 0.000 | 22:13 |
| Diagnostics Check  | OK    | 22:13 |

---

Data Download Date/Time: 05/21/2016 04:49

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RAMOS,ROBERT,G  
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)  
ARRESTING OFFICER: MCHENRY,JAMES  
ARRESTING OFFICER AGENCY: FFWCC (Panama City)  
LAST AGENCY INSPECTION: 04/13/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03415080A2  
DRY GAS EXPIRY DATE: 03/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 21:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: MIMMS

DATE/TIME OF TEST: 04/26/2016 22:27:14

---

|                   |       |       |                           |
|-------------------|-------|-------|---------------------------|
| Diagnostics Check | OK    | 22:25 |                           |
| Air Blank         | 0.000 | 22:26 |                           |
| Control Test      | 0.073 | 22:26 | Control Outside Tolerance |
| Air Blank         | 0.000 | 22:27 |                           |

---

Data Download Date/Time: 05/21/2016 04:49

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RAMOS,ROBERT,G  
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)  
ARRESTING OFFICER: MCHENRY,JAMES  
ARRESTING OFFICER AGENCY: FFWCC (Panama City)  
LAST AGENCY INSPECTION: 04/13/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03415080A2  
DRY GAS EXPIRY DATE: 03/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 21:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:55

SUBJECT'S LAST NAME: MIMMS

DATE/TIME OF TEST: 04/26/2016 22:37:56

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:31 |
| Air Blank          | 0.000 | 22:31 |
| Control Test       | 0.077 | 22:32 |
| Air Blank          | 0.000 | 22:32 |
| Subject Sample # 1 | 0.336 | 22:33 |
| Breath Volume      | 1.437 | 22:33 |
| Air Blank          | 0.000 | 22:34 |
| Air Blank          | 0.000 | 22:35 |
| Subject Sample # 2 | 0.320 | 22:36 |
| Breath Volume      | 2.394 | 22:36 |
| Air Blank          | 0.000 | 22:37 |
| Control Test       | 0.075 | 22:37 |
| Air Blank          | 0.000 | 22:37 |
| Diagnostics Check  | OK    | 22:37 |

---

Data Download Date/Time: 05/21/2016 04:49

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HUDSON,JOSHUA,K  
BREATH TEST OPERATOR AGENCY: Pensacola PD  
ARRESTING OFFICER: HUDSON,JOSHUA  
ARRESTING OFFICER AGENCY: Pensacola PD  
LAST AGENCY INSPECTION: 04/13/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03415080A2  
DRY GAS EXPIRY DATE: 03/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 01:22  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: OUELLETTE

DATE/TIME OF TEST: 04/27/2016 02:06:23

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:59 |
| Air Blank          | 0.000 | 01:59 |
| Control Test       | 0.078 | 02:00 |
| Air Blank          | 0.000 | 02:00 |
| Subject Sample # 1 | 0.196 | 02:01 |
| Breath Volume      | 2.199 | 02:01 |
| Air Blank          | 0.000 | 02:02 |
| Air Blank          | 0.000 | 02:04 |
| Subject Sample # 2 | 0.186 | 02:04 |
| Breath Volume      | 2.335 | 02:04 |
| Air Blank          | 0.000 | 02:05 |
| Control Test       | 0.079 | 02:05 |
| Air Blank          | 0.000 | 02:06 |
| Diagnostics Check  | OK    | 02:06 |

---

Data Download Date/Time: 05/21/2016 04:49

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LARMEIR,KRISTIN,D  
BREATH TEST OPERATOR AGENCY: Escambia County SO  
ARRESTING OFFICER: BURKHART,MERLE  
ARRESTING OFFICER AGENCY: Escambia County SO  
LAST AGENCY INSPECTION: 04/13/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03415080A2  
DRY GAS EXPIRY DATE: 03/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 23:05  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: LARMEIR

DATE/TIME OF TEST: 04/28/2016 00:28:37

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:22 |
| Air Blank          | 0.000 | 00:22 |
| Control Test       | 0.078 | 00:22 |
| Air Blank          | 0.000 | 00:23 |
| Subject Sample # 1 | 0.312 | 00:24 |
| Breath Volume      | 1.601 | 00:24 |
| Air Blank          | 0.000 | 00:24 |
| Air Blank          | 0.000 | 00:26 |
| Subject Sample # 2 | 0.313 | 00:27 |
| Breath Volume      | 1.964 | 00:27 |
| Air Blank          | 0.000 | 00:27 |
| Control Test       | 0.078 | 00:28 |
| Air Blank          | 0.000 | 00:28 |
| Diagnostics Check  | OK    | 00:28 |

---

Data Download Date/Time: 05/21/2016 04:49

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BUSTO,CHRISTOPHER,A  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: BUSTO,CHRISTOPHER  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 04/13/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03415080A2  
DRY GAS EXPIRY DATE: 03/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 01:38  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:38

SUBJECT'S LAST NAME: COATES

DATE/TIME OF TEST: 04/28/2016 02:13:56

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:07 |
| Air Blank          | 0.000 | 02:08 |
| Control Test       | 0.079 | 02:08 |
| Air Blank          | 0.000 | 02:08 |
| Subject Sample # 1 | 0.000 | 02:09 |
| Breath Volume      | 1.785 | 02:09 |
| Air Blank          | 0.000 | 02:09 |
| Air Blank          | 0.000 | 02:11 |
| Subject Sample # 2 | 0.000 | 02:12 |
| Breath Volume      | 1.164 | 02:12 |
| Air Blank          | 0.000 | 02:13 |
| Control Test       | 0.079 | 02:13 |
| Air Blank          | 0.000 | 02:13 |
| Diagnostics Check  | OK    | 02:13 |

---

Data Download Date/Time: 05/21/2016 04:49



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: REGISTER, MICHAEL, A  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: REGISTER, MICHAEL  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 04/13/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03415080A2  
DRY GAS EXPIRY DATE: 03/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 09:45  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 09:50

SUBJECT'S LAST NAME: BERTHIAUME

DATE/TIME OF TEST: 04/28/2016 10:18:02

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 10:10 |
| Air Blank          | 0.000 | 10:10 |
| Control Test       | 0.079 | 10:11 |
| Air Blank          | 0.000 | 10:11 |
| Subject Sample # 1 | 0.000 | 10:13 |
| Breath Volume      | 1.296 | 10:13 |
| Air Blank          | 0.000 | 10:14 |
| Air Blank          | 0.000 | 10:16 |
| Subject Sample # 2 | 0.000 | 10:16 |
| Breath Volume      | 1.246 | 10:16 |
| Air Blank          | 0.000 | 10:17 |
| Control Test       | 0.078 | 10:17 |
| Air Blank          | 0.000 | 10:17 |
| Diagnostics Check  | OK    | 10:18 |

---

Data Download Date/Time: 05/21/2016 04:50

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HUDSON,JOSHUA,K  
BREATH TEST OPERATOR AGENCY: Pensacola PD  
ARRESTING OFFICER: HUDSON,JOSHUA  
ARRESTING OFFICER AGENCY: Pensacola PD  
LAST AGENCY INSPECTION: 04/13/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03415080A2  
DRY GAS EXPIRY DATE: 03/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 01:09  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:23

SUBJECT'S LAST NAME: ALEXANDER

DATE/TIME OF TEST: 04/29/2016 02:01:03

---

|                   |       |       |                    |
|-------------------|-------|-------|--------------------|
| Diagnostics Check | OK    | 01:49 |                    |
| Air Blank         | 0.000 | 01:50 |                    |
| Control Test      | 0.078 | 01:50 |                    |
| Air Blank         | 0.000 | 01:50 |                    |
| Subject Sample    | # 1   | 01:54 | No Sample Provided |
| Breath Volume     | 0.000 | 01:54 |                    |
| Air Blank         | 0.000 | 01:54 |                    |
| Air Blank         | 0.000 | 01:56 |                    |
| Subject Sample    | # 2   | 01:59 | No Sample Provided |
| Breath Volume     | 0.000 | 01:59 |                    |
| Air Blank         | 0.000 | 02:00 |                    |
| Control Test      | 0.078 | 02:00 |                    |
| Air Blank         | 0.000 | 02:00 |                    |
| Diagnostics Check | OK    | 02:01 |                    |

---

Data Download Date/Time: 05/21/2016 04:50

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MEEKS, JEREMIAH, S  
BREATH TEST OPERATOR AGENCY: Escambia County SO  
ARRESTING OFFICER: MEEKS, JEREMIAH  
ARRESTING OFFICER AGENCY: Escambia County SO  
LAST AGENCY INSPECTION: 04/13/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03415080A2  
DRY GAS EXPIRY DATE: 03/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 22:58  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:38

SUBJECT'S LAST NAME: DIAMOND

DATE/TIME OF TEST: 04/30/2016 00:07:05

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:58 |
| Air Blank          | 0.000 | 23:59 |
| Control Test       | 0.079 | 23:59 |
| Air Blank          | 0.000 | 00:00 |
| Subject Sample # 1 | 0.080 | 00:02 |
| Breath Volume      | 3.242 | 00:02 |
| Air Blank          | 0.000 | 00:03 |
| Air Blank          | 0.000 | 00:05 |
| Subject Sample # 2 | 0.079 | 00:05 |
| Breath Volume      | 4.171 | 00:05 |
| Air Blank          | 0.000 | 00:06 |
| Control Test       | 0.077 | 00:06 |
| Air Blank          | 0.000 | 00:06 |
| Diagnostics Check  | OK    | 00:07 |

---

Data Download Date/Time: 05/21/2016 04:50

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CAIN,JEFFRY,A  
BREATH TEST OPERATOR AGENCY: Pensacola PD  
ARRESTING OFFICER: LOWRY,BRYAN  
ARRESTING OFFICER AGENCY: Pensacola PD  
LAST AGENCY INSPECTION: 04/13/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03415080A2  
DRY GAS EXPIRY DATE: 03/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 01:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: WANDS

DATE/TIME OF TEST: 04/30/2016 01:50:43

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 01:47 |                      |
| Air Blank         | 0.000 | 01:47 |                      |
| Control Test      | 0.078 | 01:47 |                      |
| Air Blank         | 0.000 | 01:48 |                      |
| Subject Sample    | # 1   | 01:49 | Subject Test Refused |
| Breath Volume     | 0.000 | 01:49 |                      |
| Air Blank         | 0.000 | 01:49 |                      |
| Control Test      | 0.078 | 01:50 |                      |
| Air Blank         | 0.000 | 01:50 |                      |
| Diagnostics Check | OK    | 01:50 |                      |

---

Data Download Date/Time: 05/21/2016 04:50

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HUDSON,JOSHUA,K  
BREATH TEST OPERATOR AGENCY: Pensacola PD  
ARRESTING OFFICER: BYRD,BRANDON  
ARRESTING OFFICER AGENCY: Pensacola PD  
LAST AGENCY INSPECTION: 04/13/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03415080A2  
DRY GAS EXPIRY DATE: 03/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 02:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:01

SUBJECT'S LAST NAME: BOATNER

DATE/TIME OF TEST: 04/30/2016 03:38:12

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 03:35 |                      |
| Air Blank         | 0.000 | 03:35 |                      |
| Control Test      | 0.078 | 03:36 |                      |
| Air Blank         | 0.000 | 03:36 |                      |
| Subject Sample    | # 1   | 03:36 | Subject Test Refused |
| Breath Volume     | 0.000 | 03:36 |                      |
| Air Blank         | 0.000 | 03:37 |                      |
| Control Test      | 0.078 | 03:37 |                      |
| Air Blank         | 0.000 | 03:38 |                      |
| Diagnostics Check | OK    | 03:38 |                      |

---

Data Download Date/Time: 05/21/2016 04:50

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GORMLEY, WILLIAM, C  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: GORMLEY, WILLIAM  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 04/13/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03415080A2  
DRY GAS EXPIRY DATE: 03/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 13:09  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 13:39

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 04/30/2016 14:08:44

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 13:59 |
| Air Blank          | 0.000 | 14:00 |
| Control Test       | 0.079 | 14:00 |
| Air Blank          | 0.000 | 14:00 |
| Subject Sample # 1 | 0.331 | 14:02 |
| Breath Volume      | 1.855 | 14:02 |
| Air Blank          | 0.000 | 14:03 |
| Air Blank          | 0.000 | 14:04 |
| Subject Sample # 2 | 0.325 | 14:06 |
| Breath Volume      | 1.671 | 14:06 |
| Air Blank          | 0.000 | 14:07 |
| Control Test       | 0.076 | 14:08 |
| Air Blank          | 0.000 | 14:08 |
| Diagnostics Check  | OK    | 14:08 |

---

Data Download Date/Time: 05/21/2016 04:50

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MEEKS, JEREMIAH, S  
BREATH TEST OPERATOR AGENCY: Escambia County SO  
ARRESTING OFFICER: MEEKS, JEREMIAH  
ARRESTING OFFICER AGENCY: Escambia County SO  
LAST AGENCY INSPECTION: 04/13/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03415080A2  
DRY GAS EXPIRY DATE: 03/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 17:27  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:08

SUBJECT'S LAST NAME: PORTER

DATE/TIME OF TEST: 04/30/2016 18:35:19

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 18:29 |
| Air Blank          | 0.000 | 18:29 |
| Control Test       | 0.078 | 18:29 |
| Air Blank          | 0.000 | 18:30 |
| Subject Sample # 1 | 0.034 | 18:30 |
| Breath Volume      | 3.253 | 18:30 |
| Air Blank          | 0.000 | 18:31 |
| Air Blank          | 0.000 | 18:33 |
| Subject Sample # 2 | 0.034 | 18:33 |
| Breath Volume      | 3.324 | 18:33 |
| Air Blank          | 0.000 | 18:34 |
| Control Test       | 0.077 | 18:34 |
| Air Blank          | 0.000 | 18:35 |
| Diagnostics Check  | OK    | 18:35 |

---

Data Download Date/Time: 05/21/2016 04:50

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000759

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CARLBERG,DAVID,L  
BREATH TEST OPERATOR AGENCY: Jackson County SO  
ARRESTING OFFICER: MITCHELL,WESLEY  
ARRESTING OFFICER AGENCY: Jackson County SO  
LAST AGENCY INSPECTION: 03/23/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19615080A3  
DRY GAS EXPIRY DATE: 09/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 21:02  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:15

SUBJECT'S LAST NAME: GRAY

DATE/TIME OF TEST: 04/15/2016 22:50:43

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:44 |
| Air Blank          | 0.000 | 22:44 |
| Control Test       | 0.077 | 22:45 |
| Air Blank          | 0.000 | 22:45 |
| Subject Sample # 1 | 0.223 | 22:46 |
| Breath Volume      | 2.640 | 22:46 |
| Air Blank          | 0.000 | 22:46 |
| Air Blank          | 0.000 | 22:48 |
| Subject Sample # 2 | 0.223 | 22:49 |
| Breath Volume      | 2.480 | 22:49 |
| Air Blank          | 0.000 | 22:49 |
| Control Test       | 0.075 | 22:50 |
| Air Blank          | 0.000 | 22:50 |
| Diagnostics Check  | OK    | 22:50 |

---

Data Download Date/Time: 04/23/2016 21:17



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HUNT,CHARLES,R  
BREATH TEST OPERATOR AGENCY: Alachua County SO  
ARRESTING OFFICER: HUNT,CHARLES  
ARRESTING OFFICER AGENCY: Alachua County SO  
LAST AGENCY INSPECTION: 03/25/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 11315080A1  
DRY GAS EXPIRY DATE: 05/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 03:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:07

SUBJECT'S LAST NAME: CARABEO

DATE/TIME OF TEST: 04/01/2016 04:44:09

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 04:34 |                      |
| Air Blank         | 0.000 | 04:35 |                      |
| Control Test      | 0.078 | 04:35 |                      |
| Air Blank         | 0.000 | 04:35 |                      |
| Subject Sample    | # 1   | 04:39 | No Sample Provided   |
| Breath Volume     | 0.000 | 04:39 |                      |
| Air Blank         | 0.000 | 04:39 |                      |
| Air Blank         | 0.000 | 04:41 |                      |
| Subject Sample    | # 2   | 04:42 | Subject Test Refused |
| Breath Volume     | 0.000 | 04:42 |                      |
| Air Blank         | 0.000 | 04:43 |                      |
| Control Test      | 0.079 | 04:43 |                      |
| Air Blank         | 0.000 | 04:44 |                      |
| Diagnostics Check | OK    | 04:44 |                      |

---

Data Download Date/Time: 04/26/2016 09:04

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PALLETT,LANCE,P  
BREATH TEST OPERATOR AGENCY: Alachua County SO  
ARRESTING OFFICER: PALLETT,LANCE  
ARRESTING OFFICER AGENCY: Alachua County SO  
LAST AGENCY INSPECTION: 03/25/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 11315080A1  
DRY GAS EXPIRY DATE: 05/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 20:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:38

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 04/01/2016 22:11:05

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 22:08 |                      |
| Air Blank         | 0.000 | 22:08 |                      |
| Control Test      | 0.079 | 22:08 |                      |
| Air Blank         | 0.000 | 22:09 |                      |
| Subject Sample    | # 1   | 22:09 | Subject Test Refused |
| Breath Volume     | 0.000 | 22:09 |                      |
| Air Blank         | 0.000 | 22:10 |                      |
| Control Test      | 0.078 | 22:10 |                      |
| Air Blank         | 0.000 | 22:11 |                      |
| Diagnostics Check | OK    | 22:11 |                      |

---

Data Download Date/Time: 04/26/2016 09:04

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: QUALMANN,DANIEL,G  
BREATH TEST OPERATOR AGENCY: Alachua County SO  
ARRESTING OFFICER: QUALMANN,DANIEL  
ARRESTING OFFICER AGENCY: Alachua County SO  
LAST AGENCY INSPECTION: 03/25/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 11315080A1  
DRY GAS EXPIRY DATE: 05/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 23:03  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:16

SUBJECT'S LAST NAME: HILBRAND

DATE/TIME OF TEST: 04/02/2016 00:50:16

---

|                   |           |       |                    |
|-------------------|-----------|-------|--------------------|
| Diagnostics Check | OK        | 00:38 |                    |
| Air Blank         | 0.000     | 00:39 |                    |
| Control Test      | 0.078     | 00:39 |                    |
| Air Blank         | 0.000     | 00:40 |                    |
| Subject Sample    | # 1 0.124 | 00:43 | Volume Not Met     |
| Breath Volume     | 0.058     | 00:43 |                    |
| Air Blank         | 0.000     | 00:44 |                    |
| Air Blank         | 0.000     | 00:45 |                    |
| Subject Sample    | # 2       | 00:48 | No Sample Provided |
| Breath Volume     | 0.000     | 00:48 |                    |
| Air Blank         | 0.000     | 00:49 |                    |
| Control Test      | 0.078     | 00:49 |                    |
| Air Blank         | 0.000     | 00:50 |                    |
| Diagnostics Check | OK        | 00:50 |                    |

---

Data Download Date/Time: 04/26/2016 09:04

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JONES,THEODORE,D  
BREATH TEST OPERATOR AGENCY: Gainesville PD  
ARRESTING OFFICER: BOHANNON,MICHAEL  
ARRESTING OFFICER AGENCY: Alachua PD  
LAST AGENCY INSPECTION: 03/25/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 11315080A1  
DRY GAS EXPIRY DATE: 05/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 01:28  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:23

SUBJECT'S LAST NAME: FREEMAN

DATE/TIME OF TEST: 04/02/2016 02:49:56

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:43 |
| Air Blank          | 0.000 | 02:44 |
| Control Test       | 0.078 | 02:44 |
| Air Blank          | 0.000 | 02:44 |
| Subject Sample # 1 | 0.146 | 02:45 |
| Breath Volume      | 2.320 | 02:45 |
| Air Blank          | 0.000 | 02:46 |
| Air Blank          | 0.000 | 02:47 |
| Subject Sample # 2 | 0.152 | 02:48 |
| Breath Volume      | 2.929 | 02:48 |
| Air Blank          | 0.000 | 02:49 |
| Control Test       | 0.076 | 02:49 |
| Air Blank          | 0.000 | 02:49 |
| Diagnostics Check  | OK    | 02:49 |

---

Data Download Date/Time: 04/26/2016 09:04

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RICHARDSON, MICHAEL, R  
BREATH TEST OPERATOR AGENCY: Alachua County SO  
ARRESTING OFFICER: STEPHENSON, JASPER  
ARRESTING OFFICER AGENCY: Alachua County SO  
LAST AGENCY INSPECTION: 03/25/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 11315080A1  
DRY GAS EXPIRY DATE: 05/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 02:37  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:24

SUBJECT'S LAST NAME: LIMA

DATE/TIME OF TEST: 04/03/2016 03:52:24

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:46 |
| Air Blank          | 0.000 | 03:46 |
| Control Test       | 0.079 | 03:46 |
| Air Blank          | 0.000 | 03:47 |
| Subject Sample # 1 | 0.193 | 03:47 |
| Breath Volume      | 1.433 | 03:47 |
| Air Blank          | 0.000 | 03:48 |
| Air Blank          | 0.000 | 03:50 |
| Subject Sample # 2 | 0.188 | 03:50 |
| Breath Volume      | 1.429 | 03:50 |
| Air Blank          | 0.000 | 03:51 |
| Control Test       | 0.080 | 03:51 |
| Air Blank          | 0.000 | 03:52 |
| Diagnostics Check  | OK    | 03:52 |

---

Data Download Date/Time: 04/26/2016 09:04

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HUNT,CHARLES,R  
BREATH TEST OPERATOR AGENCY: Alachua County SO  
ARRESTING OFFICER: SPERRING,JOSEPH  
ARRESTING OFFICER AGENCY: Alachua County SO  
LAST AGENCY INSPECTION: 03/25/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 11315080A1  
DRY GAS EXPIRY DATE: 05/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 01:49  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: MCCOY

DATE/TIME OF TEST: 04/06/2016 03:17:54

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 03:07 |                      |
| Air Blank         | 0.000 | 03:08 |                      |
| Control Test      | 0.078 | 03:08 |                      |
| Air Blank         | 0.000 | 03:09 |                      |
| Subject Sample    | # 1   | 03:12 | No Sample Provided   |
| Breath Volume     | 0.000 | 03:12 |                      |
| Air Blank         | 0.000 | 03:12 |                      |
| Air Blank         | 0.000 | 03:14 |                      |
| Subject Sample    | # 2   | 03:16 | Subject Test Refused |
| Breath Volume     | 0.000 | 03:16 |                      |
| Air Blank         | 0.000 | 03:16 |                      |
| Control Test      | 0.079 | 03:17 |                      |
| Air Blank         | 0.000 | 03:17 |                      |
| Diagnostics Check | OK    | 03:17 |                      |

---

Data Download Date/Time: 04/26/2016 09:04

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: QUALMANN,DANIEL,G  
BREATH TEST OPERATOR AGENCY: Alachua County SO  
ARRESTING OFFICER: DRAKE,CHARLES  
ARRESTING OFFICER AGENCY: Alachua County SO  
LAST AGENCY INSPECTION: 03/25/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 11315080A1  
DRY GAS EXPIRY DATE: 05/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 00:32  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: MCCOOK

DATE/TIME OF TEST: 04/07/2016 01:35:19

---

|                   |       |       |            |
|-------------------|-------|-------|------------|
| Diagnostics Check | OK    | 01:31 |            |
| Air Blank         | 0.000 | 01:31 |            |
| Control Test      | 0.079 | 01:31 |            |
| Air Blank         | 0.000 | 01:32 |            |
| Subject Sample    | # 1   | 01:33 | RFI Detect |
| Breath Volume     | 0.000 | 01:33 |            |
| Air Blank         | 0.000 | 01:34 |            |
| Control Test      | 0.079 | 01:34 |            |
| Air Blank         | 0.000 | 01:35 |            |
| Diagnostics Check | OK    | 01:35 |            |

---

Data Download Date/Time: 04/26/2016 09:04

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: QUALMANN,DANIEL,G  
BREATH TEST OPERATOR AGENCY: Alachua County SO  
ARRESTING OFFICER: DRAKE,CHARLES  
ARRESTING OFFICER AGENCY: Alachua County SO  
LAST AGENCY INSPECTION: 03/25/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 11315080A1  
DRY GAS EXPIRY DATE: 05/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 00:32  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: MCCOOK

DATE/TIME OF TEST: 04/07/2016 01:47:56

---

|                   |       |       |                    |
|-------------------|-------|-------|--------------------|
| Diagnostics Check | OK    | 01:36 |                    |
| Air Blank         | 0.000 | 01:37 |                    |
| Control Test      | 0.079 | 01:37 |                    |
| Air Blank         | 0.000 | 01:37 |                    |
| Subject Sample    | # 1   | 01:41 | No Sample Provided |
| Breath Volume     | 0.000 | 01:41 |                    |
| Air Blank         | 0.000 | 01:41 |                    |
| Air Blank         | 0.000 | 01:43 |                    |
| Subject Sample    | # 2   | 01:46 | No Sample Provided |
| Breath Volume     | 0.000 | 01:46 |                    |
| Air Blank         | 0.000 | 01:47 |                    |
| Control Test      | 0.079 | 01:47 |                    |
| Air Blank         | 0.000 | 01:47 |                    |
| Diagnostics Check | OK    | 01:47 |                    |

---

Data Download Date/Time: 04/26/2016 09:04



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BOSTICK,JESSE,W  
BREATH TEST OPERATOR AGENCY: Gainesville PD  
ARRESTING OFFICER: BOSTICK,JESSE  
ARRESTING OFFICER AGENCY: Gainesville PD  
LAST AGENCY INSPECTION: 03/25/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 11315080A1  
DRY GAS EXPIRY DATE: 05/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 05:26  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:11

SUBJECT'S LAST NAME: RAWLS JR

DATE/TIME OF TEST: 04/07/2016 06:40:52

---

|                   |       |       |            |
|-------------------|-------|-------|------------|
| Diagnostics Check | OK    | 06:37 |            |
| Air Blank         | 0.000 | 06:37 |            |
| Control Test      | 0.078 | 06:38 |            |
| Air Blank         | 0.000 | 06:38 |            |
| Subject Sample    | # 1   | 06:39 | RFI Detect |
| Breath Volume     | 0.890 | 06:39 |            |
| Air Blank         | 0.000 | 06:39 |            |
| Control Test      | 0.079 | 06:40 |            |
| Air Blank         | 0.000 | 06:40 |            |
| Diagnostics Check | OK    | 06:40 |            |

---

Data Download Date/Time: 04/26/2016 09:04

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BOSTICK,JESSE,W  
BREATH TEST OPERATOR AGENCY: Gainesville PD  
ARRESTING OFFICER: BOSTICK,JESSE  
ARRESTING OFFICER AGENCY: Gainesville PD  
LAST AGENCY INSPECTION: 03/25/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 11315080A1  
DRY GAS EXPIRY DATE: 05/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 05:26  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:11

SUBJECT'S LAST NAME: RAWLS JR

DATE/TIME OF TEST: 04/07/2016 06:45:13

---

|                   |       |       |            |
|-------------------|-------|-------|------------|
| Diagnostics Check | OK    | 06:42 |            |
| Air Blank         | 0.000 | 06:42 |            |
| Control Test      | 0.079 | 06:43 |            |
| Air Blank         | 0.000 | 06:43 |            |
| Subject Sample    | # 1   | 06:43 | RFI Detect |
| Breath Volume     | 0.000 | 06:43 |            |
| Air Blank         | 0.000 | 06:44 |            |
| Control Test      | 0.078 | 06:44 |            |
| Air Blank         | 0.000 | 06:45 |            |
| Diagnostics Check | OK    | 06:45 |            |

---

Data Download Date/Time: 04/26/2016 09:04

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BOSTICK,JESSE,W  
BREATH TEST OPERATOR AGENCY: Gainesville PD  
ARRESTING OFFICER: BOSTICK,JESSE  
ARRESTING OFFICER AGENCY: Gainesville PD  
LAST AGENCY INSPECTION: 03/25/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 11315080A1  
DRY GAS EXPIRY DATE: 05/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 05:26  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:11

SUBJECT'S LAST NAME: RAWLS JR

DATE/TIME OF TEST: 04/07/2016 06:53:25

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 06:46 |
| Air Blank          | 0.000 | 06:47 |
| Control Test       | 0.078 | 06:47 |
| Air Blank          | 0.000 | 06:48 |
| Subject Sample # 1 | 0.122 | 06:48 |
| Breath Volume      | 1.582 | 06:48 |
| Air Blank          | 0.000 | 06:49 |
| Air Blank          | 0.000 | 06:51 |
| Subject Sample # 2 | 0.113 | 06:51 |
| Breath Volume      | 1.921 | 06:51 |
| Air Blank          | 0.000 | 06:52 |
| Control Test       | 0.078 | 06:52 |
| Air Blank          | 0.000 | 06:53 |
| Diagnostics Check  | OK    | 06:53 |

---

Data Download Date/Time: 04/26/2016 09:04

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LANDER,JAMES,P  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: LANDER,JAMES  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 03/25/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 11315080A1  
DRY GAS EXPIRY DATE: 05/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 12:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 13:48

SUBJECT'S LAST NAME: ROBERTS

DATE/TIME OF TEST: 04/08/2016 14:15:26

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 14:09 |
| Air Blank          | 0.000 | 14:09 |
| Control Test       | 0.079 | 14:09 |
| Air Blank          | 0.000 | 14:10 |
| Subject Sample # 1 | 0.263 | 14:10 |
| Breath Volume      | 2.851 | 14:10 |
| Air Blank          | 0.000 | 14:11 |
| Air Blank          | 0.000 | 14:13 |
| Subject Sample # 2 | 0.246 | 14:13 |
| Breath Volume      | 2.414 | 14:13 |
| Air Blank          | 0.000 | 14:14 |
| Control Test       | 0.077 | 14:14 |
| Air Blank          | 0.000 | 14:15 |
| Diagnostics Check  | OK    | 14:15 |

---

Data Download Date/Time: 04/26/2016 09:04

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HUNT,CHARLES,R  
BREATH TEST OPERATOR AGENCY: Alachua County SO  
ARRESTING OFFICER: HUNT,CHARLES  
ARRESTING OFFICER AGENCY: Alachua County SO  
LAST AGENCY INSPECTION: 03/25/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 11315080A1  
DRY GAS EXPIRY DATE: 05/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 22:03  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:47

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 04/08/2016 23:22:14

---

|                   |           |       |                      |
|-------------------|-----------|-------|----------------------|
| Diagnostics Check | OK        | 23:11 |                      |
| Air Blank         | 0.000     | 23:12 |                      |
| Control Test      | 0.079     | 23:12 |                      |
| Air Blank         | 0.000     | 23:12 |                      |
| Subject Sample    | # 1 0.201 | 23:15 | Volume Not Met       |
| Breath Volume     | 0.464     | 23:15 |                      |
| Air Blank         | 0.000     | 23:16 |                      |
| Air Blank         | 0.000     | 23:18 |                      |
| Subject Sample    | # 2       | 23:20 | Subject Test Refused |
| Breath Volume     | 0.132     | 23:20 |                      |
| Air Blank         | 0.000     | 23:21 |                      |
| Control Test      | 0.076     | 23:21 |                      |
| Air Blank         | 0.000     | 23:22 |                      |
| Diagnostics Check | OK        | 23:22 |                      |

---

Data Download Date/Time: 04/26/2016 09:04

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HOFMANN, TRAVIS, A  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: CAULK, DANIEL  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 03/25/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 11315080A1  
DRY GAS EXPIRY DATE: 05/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 23:56  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: CARR

DATE/TIME OF TEST: 04/15/2016 01:52:37

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:45 |
| Air Blank          | 0.000 | 01:46 |
| Control Test       | 0.079 | 01:46 |
| Air Blank          | 0.000 | 01:47 |
| Subject Sample # 1 | 0.199 | 01:48 |
| Breath Volume      | 1.796 | 01:48 |
| Air Blank          | 0.000 | 01:48 |
| Air Blank          | 0.000 | 01:50 |
| Subject Sample # 2 | 0.207 | 01:51 |
| Breath Volume      | 2.121 | 01:51 |
| Air Blank          | 0.000 | 01:51 |
| Control Test       | 0.079 | 01:52 |
| Air Blank          | 0.000 | 01:52 |
| Diagnostics Check  | OK    | 01:52 |

---

Data Download Date/Time: 04/26/2016 09:04

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HUNT,CHARLES,R  
BREATH TEST OPERATOR AGENCY: Alachua County SO  
ARRESTING OFFICER: HUNT,CHARLES  
ARRESTING OFFICER AGENCY: Alachua County SO  
LAST AGENCY INSPECTION: 03/25/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 11315080A1  
DRY GAS EXPIRY DATE: 05/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 01:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:08

SUBJECT'S LAST NAME: STECHMILLER

DATE/TIME OF TEST: 04/15/2016 02:41:32

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:35 |
| Air Blank          | 0.000 | 02:35 |
| Control Test       | 0.078 | 02:35 |
| Air Blank          | 0.000 | 02:36 |
| Subject Sample # 1 | 0.000 | 02:37 |
| Breath Volume      | 1.660 | 02:37 |
| Air Blank          | 0.000 | 02:37 |
| Air Blank          | 0.000 | 02:39 |
| Subject Sample # 2 | 0.000 | 02:40 |
| Breath Volume      | 1.269 | 02:40 |
| Air Blank          | 0.000 | 02:40 |
| Control Test       | 0.078 | 02:40 |
| Air Blank          | 0.000 | 02:41 |
| Diagnostics Check  | OK    | 02:41 |

---

Data Download Date/Time: 04/26/2016 09:04

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PRIVETTE,CHRIS,A  
BREATH TEST OPERATOR AGENCY: Alachua County SO  
ARRESTING OFFICER: PRIVETTE,CHRIS  
ARRESTING OFFICER AGENCY: Alachua County SO  
LAST AGENCY INSPECTION: 03/25/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 11315080A1  
DRY GAS EXPIRY DATE: 05/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 11:40  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 12:37

SUBJECT'S LAST NAME: SULLIVAN

DATE/TIME OF TEST: 04/15/2016 13:11:08

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 13:04 |
| Air Blank          | 0.000 | 13:05 |
| Control Test       | 0.079 | 13:05 |
| Air Blank          | 0.000 | 13:06 |
| Subject Sample # 1 | 0.000 | 13:06 |
| Breath Volume      | 2.277 | 13:06 |
| Air Blank          | 0.000 | 13:07 |
| Air Blank          | 0.000 | 13:09 |
| Subject Sample # 2 | 0.000 | 13:09 |
| Breath Volume      | 2.187 | 13:09 |
| Air Blank          | 0.000 | 13:10 |
| Control Test       | 0.079 | 13:10 |
| Air Blank          | 0.000 | 13:11 |
| Diagnostics Check  | OK    | 13:11 |

---

Data Download Date/Time: 04/26/2016 09:04



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: QUALMANN,DANIEL,G  
BREATH TEST OPERATOR AGENCY: Alachua County SO  
ARRESTING OFFICER: QUALMANN,DANIEL  
ARRESTING OFFICER AGENCY: Alachua County SO  
LAST AGENCY INSPECTION: 03/25/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 11315080A1  
DRY GAS EXPIRY DATE: 05/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 00:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: CAMPBELL

DATE/TIME OF TEST: 04/17/2016 02:14:41

---

|                   |       |       |                    |
|-------------------|-------|-------|--------------------|
| Diagnostics Check | OK    | 02:03 |                    |
| Air Blank         | 0.000 | 02:03 |                    |
| Control Test      | 0.079 | 02:04 |                    |
| Air Blank         | 0.000 | 02:04 |                    |
| Subject Sample    | # 1   | 02:07 | No Sample Provided |
| Breath Volume     | 0.000 | 02:07 |                    |
| Air Blank         | 0.000 | 02:08 |                    |
| Air Blank         | 0.000 | 02:10 |                    |
| Subject Sample    | # 2   | 02:13 | No Sample Provided |
| Breath Volume     | 0.000 | 02:13 |                    |
| Air Blank         | 0.000 | 02:13 |                    |
| Control Test      | 0.080 | 02:14 |                    |
| Air Blank         | 0.000 | 02:14 |                    |
| Diagnostics Check | OK    | 02:14 |                    |

---

Data Download Date/Time: 04/26/2016 09:04

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: QUALMANN,DANIEL,G  
BREATH TEST OPERATOR AGENCY: Alachua County SO  
ARRESTING OFFICER: STEPHENSON,JASPER  
ARRESTING OFFICER AGENCY: Alachua County SO  
LAST AGENCY INSPECTION: 03/25/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 11315080A1  
DRY GAS EXPIRY DATE: 05/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 04:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:57

SUBJECT'S LAST NAME: JERNIGAN

DATE/TIME OF TEST: 04/17/2016 05:27:08

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 05:19 |
| Air Blank          | 0.000 | 05:20 |
| Control Test       | 0.078 | 05:20 |
| Air Blank          | 0.000 | 05:21 |
| Subject Sample # 1 | 0.071 | 05:22 |
| Breath Volume      | 1.250 | 05:22 |
| Air Blank          | 0.000 | 05:22 |
| Air Blank          | 0.000 | 05:24 |
| Subject Sample # 2 | 0.075 | 05:25 |
| Breath Volume      | 4.218 | 05:25 |
| Air Blank          | 0.000 | 05:26 |
| Control Test       | 0.078 | 05:26 |
| Air Blank          | 0.000 | 05:27 |
| Diagnostics Check  | OK    | 05:27 |

---

Data Download Date/Time: 04/26/2016 09:04

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MCFARLAND,MICHAEL,D  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: MCFARLAND,MICHAEL  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 03/25/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 11315080A1  
DRY GAS EXPIRY DATE: 05/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 22:56  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: KARRAS

DATE/TIME OF TEST: 04/22/2016 00:51:53

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:45 |
| Air Blank          | 0.000 | 00:45 |
| Control Test       | 0.079 | 00:46 |
| Air Blank          | 0.000 | 00:46 |
| Subject Sample # 1 | 0.140 | 00:47 |
| Breath Volume      | 3.722 | 00:47 |
| Air Blank          | 0.000 | 00:48 |
| Air Blank          | 0.000 | 00:49 |
| Subject Sample # 2 | 0.127 | 00:50 |
| Breath Volume      | 1.699 | 00:50 |
| Air Blank          | 0.000 | 00:50 |
| Control Test       | 0.078 | 00:51 |
| Air Blank          | 0.000 | 00:51 |
| Diagnostics Check  | OK    | 00:51 |

---

Data Download Date/Time: 04/26/2016 09:04

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BOSTICK,JESSE,W  
BREATH TEST OPERATOR AGENCY: Gainesville PD  
ARRESTING OFFICER: WENDT,HENDRIK  
ARRESTING OFFICER AGENCY: Gainesville PD  
LAST AGENCY INSPECTION: 03/25/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 11315080A1  
DRY GAS EXPIRY DATE: 05/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 03:48  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:03

SUBJECT'S LAST NAME: MORABITO

DATE/TIME OF TEST: 04/22/2016 04:28:23

---

|                   |       |       |            |
|-------------------|-------|-------|------------|
| Diagnostics Check | OK    | 04:27 |            |
| Air Blank         |       | 04:27 | RFI Detect |
| Air Blank         | 0.000 | 04:28 |            |

---

Data Download Date/Time: 04/26/2016 09:04

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BOSTICK,JESSE,W  
BREATH TEST OPERATOR AGENCY: Gainesville PD  
ARRESTING OFFICER: WENDT,HENDRIK  
ARRESTING OFFICER AGENCY: Gainesville PD  
LAST AGENCY INSPECTION: 03/25/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 11315080A1  
DRY GAS EXPIRY DATE: 05/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 03:48  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:03

SUBJECT'S LAST NAME: MORABITO

DATE/TIME OF TEST: 04/22/2016 04:36:20

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:30 |
| Air Blank          | 0.000 | 04:30 |
| Control Test       | 0.079 | 04:31 |
| Air Blank          | 0.000 | 04:31 |
| Subject Sample # 1 | 0.000 | 04:32 |
| Breath Volume      | 1.816 | 04:32 |
| Air Blank          | 0.000 | 04:32 |
| Air Blank          | 0.000 | 04:34 |
| Subject Sample # 2 | 0.000 | 04:34 |
| Breath Volume      | 1.554 | 04:34 |
| Air Blank          | 0.000 | 04:35 |
| Control Test       | 0.079 | 04:35 |
| Air Blank          | 0.000 | 04:36 |
| Diagnostics Check  | OK    | 04:36 |

---

Data Download Date/Time: 04/26/2016 09:04

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HUNT,CHARLES,R  
BREATH TEST OPERATOR AGENCY: Alachua County SO  
ARRESTING OFFICER: HUNT,CHARLES  
ARRESTING OFFICER AGENCY: Alachua County SO  
LAST AGENCY INSPECTION: 03/25/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 11315080A1  
DRY GAS EXPIRY DATE: 05/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 03:02  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:52

SUBJECT'S LAST NAME: OWENS

DATE/TIME OF TEST: 04/24/2016 04:27:50

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:21 |
| Air Blank          | 0.000 | 04:22 |
| Control Test       | 0.079 | 04:22 |
| Air Blank          | 0.000 | 04:22 |
| Subject Sample # 1 | 0.100 | 04:23 |
| Breath Volume      | 1.171 | 04:23 |
| Air Blank          | 0.000 | 04:23 |
| Air Blank          | 0.000 | 04:25 |
| Subject Sample # 2 | 0.094 | 04:26 |
| Breath Volume      | 1.117 | 04:26 |
| Air Blank          | 0.000 | 04:26 |
| Control Test       | 0.078 | 04:27 |
| Air Blank          | 0.000 | 04:27 |
| Diagnostics Check  | OK    | 04:27 |

---

Data Download Date/Time: 04/26/2016 09:04

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HOFMANN, TRAVIS, A  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: HOFMANN, TRAVIS  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 04/26/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 11315080A1  
DRY GAS EXPIRY DATE: 05/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 02:45  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:39

SUBJECT'S LAST NAME: RATLIFF

DATE/TIME OF TEST: 04/28/2016 04:12:58

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:05 |
| Air Blank          | 0.000 | 04:06 |
| Control Test       | 0.079 | 04:06 |
| Air Blank          | 0.000 | 04:07 |
| Subject Sample # 1 | 0.070 | 04:08 |
| Breath Volume      | 1.769 | 04:08 |
| Air Blank          | 0.000 | 04:08 |
| Air Blank          | 0.000 | 04:10 |
| Subject Sample # 2 | 0.065 | 04:11 |
| Breath Volume      | 1.234 | 04:11 |
| Air Blank          | 0.000 | 04:12 |
| Control Test       | 0.079 | 04:12 |
| Air Blank          | 0.000 | 04:12 |
| Diagnostics Check  | OK    | 04:12 |

---

Data Download Date/Time: 05/25/2016 08:18

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FRICKER,JEFFREY,S  
BREATH TEST OPERATOR AGENCY: Alachua County SO  
ARRESTING OFFICER: FRICKER,JEFFREY  
ARRESTING OFFICER AGENCY: Alachua County SO  
LAST AGENCY INSPECTION: 04/26/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 11315080A1  
DRY GAS EXPIRY DATE: 05/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 17:44  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:24

SUBJECT'S LAST NAME: SWALLOW

DATE/TIME OF TEST: 04/30/2016 19:00:50

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 18:53 |
| Air Blank          | 0.000 | 18:54 |
| Control Test       | 0.078 | 18:54 |
| Air Blank          | 0.000 | 18:55 |
| Subject Sample # 1 | 0.253 | 18:56 |
| Breath Volume      | 3.300 | 18:56 |
| Air Blank          | 0.000 | 18:56 |
| Air Blank          | 0.000 | 18:58 |
| Subject Sample # 2 | 0.244 | 18:59 |
| Breath Volume      | 3.621 | 18:59 |
| Air Blank          | 0.000 | 18:59 |
| Control Test       | 0.079 | 19:00 |
| Air Blank          | 0.000 | 19:00 |
| Diagnostics Check  | OK    | 19:00 |

---

Data Download Date/Time: 05/25/2016 08:18



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000770

INSTRUMENT REGISTERED TO: HAMILTON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARVEY, DAVID, A  
BREATH TEST OPERATOR AGENCY: Jasper PD  
ARRESTING OFFICER: HARVEY, DAVID  
ARRESTING OFFICER AGENCY: Jasper PD  
LAST AGENCY INSPECTION: 03/31/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06415080A2  
DRY GAS EXPIRY DATE: 04/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 00:23  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:34

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 04/17/2016 01:03:42

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 00:55 |                      |
| Air Blank         | 0.000 | 00:55 |                      |
| Control Test      | 0.078 | 00:56 |                      |
| Air Blank         | 0.000 | 00:56 |                      |
| Subject Sample    | # 1   | 00:59 | No Sample Provided   |
| Breath Volume     | 0.000 | 00:59 |                      |
| Air Blank         | 0.000 | 01:00 |                      |
| Air Blank         | 0.000 | 01:02 |                      |
| Subject Sample    | # 2   | 01:02 | Subject Test Refused |
| Breath Volume     | 0.000 | 01:02 |                      |
| Air Blank         | 0.000 | 01:02 |                      |
| Control Test      | 0.079 | 01:03 |                      |
| Air Blank         | 0.000 | 01:03 |                      |
| Diagnostics Check | OK    | 01:03 |                      |

---

Data Download Date/Time: 04/24/2016 18:09

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000770

INSTRUMENT REGISTERED TO: HAMILTON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CLARK,J,T  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: CLARK,J  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 03/31/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06415080A2  
DRY GAS EXPIRY DATE: 04/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 19:35  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:15

SUBJECT'S LAST NAME: TRIPPE

DATE/TIME OF TEST: 04/17/2016 21:49:46

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 21:46 |                      |
| Air Blank         | 0.000 | 21:47 |                      |
| Control Test      | 0.078 | 21:47 |                      |
| Air Blank         | 0.000 | 21:48 |                      |
| Subject Sample    | # 1   | 21:48 | Subject Test Refused |
| Breath Volume     | 0.000 | 21:48 |                      |
| Air Blank         | 0.000 | 21:48 |                      |
| Control Test      | 0.079 | 21:49 |                      |
| Air Blank         | 0.000 | 21:49 |                      |
| Diagnostics Check | OK    | 21:49 |                      |

---

Data Download Date/Time: 04/24/2016 18:09

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RUSSAK,KEVIN,M  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: RUSSAK,KEVIN  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 04/30/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG431502  
DRY GAS EXPIRY DATE: 11/11/2016  
VIOLATION CODE: DUI  
ARREST TIME: 22:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:31

SUBJECT'S LAST NAME: COCHENOUR

DATE/TIME OF TEST: 04/30/2016 23:36:01

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:29 |
| Air Blank          | 0.000 | 23:30 |
| Control Test       | 0.079 | 23:30 |
| Air Blank          | 0.000 | 23:31 |
| Subject Sample # 1 | 0.184 | 23:31 |
| Breath Volume      | 3.085 | 23:31 |
| Air Blank          | 0.000 | 23:32 |
| Air Blank          | 0.000 | 23:34 |
| Subject Sample # 2 | 0.185 | 23:34 |
| Breath Volume      | 3.613 | 23:34 |
| Air Blank          | 0.000 | 23:35 |
| Control Test       | 0.079 | 23:35 |
| Air Blank          | 0.000 | 23:35 |
| Diagnostics Check  | OK    | 23:36 |

---

Data Download Date/Time: 05/31/2016 15:09

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000773

INSTRUMENT REGISTERED TO: LAFAYETTE CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DEAS,ALFRED,A  
BREATH TEST OPERATOR AGENCY: Lafayette County SO  
ARRESTING OFFICER: WINBURN,ROBERT  
ARRESTING OFFICER AGENCY: Lafayette County SO  
LAST AGENCY INSPECTION: 03/28/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22315080A2  
DRY GAS EXPIRY DATE: 09/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 17:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:55

SUBJECT'S LAST NAME: GARCIA-MENDOZA

DATE/TIME OF TEST: 04/08/2016 18:30:04

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 18:23 |
| Air Blank          | 0.000 | 18:23 |
| Control Test       | 0.079 | 18:24 |
| Air Blank          | 0.000 | 18:24 |
| Subject Sample # 1 | 0.291 | 18:25 |
| Breath Volume      | 1.953 | 18:25 |
| Air Blank          | 0.000 | 18:26 |
| Air Blank          | 0.000 | 18:27 |
| Subject Sample # 2 | 0.282 | 18:28 |
| Breath Volume      | 1.667 | 18:28 |
| Air Blank          | 0.000 | 18:29 |
| Control Test       | 0.079 | 18:29 |
| Air Blank          | 0.000 | 18:29 |
| Diagnostics Check  | OK    | 18:30 |

---

Data Download Date/Time: 05/31/2016 12:58

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HIGHTOWER,CONNIE,A  
BREATH TEST OPERATOR AGENCY: Lake City PD  
ARRESTING OFFICER: MICHAELS,P  
ARRESTING OFFICER AGENCY: Lake City PD  
LAST AGENCY INSPECTION: 03/24/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1686772  
DRY GAS EXPIRY DATE: 06/02/2017  
VIOLATION CODE: DUI  
ARREST TIME: 17:58  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:00

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 04/04/2016 18:31:10

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 18:27 |                      |
| Air Blank         | 0.000 | 18:28 |                      |
| Control Test      | 0.083 | 18:28 |                      |
| Air Blank         | 0.000 | 18:29 |                      |
| Subject Sample    | # 1   | 18:29 | Subject Test Refused |
| Breath Volume     | 0.000 | 18:29 |                      |
| Air Blank         | 0.000 | 18:30 |                      |
| Control Test      | 0.082 | 18:30 |                      |
| Air Blank         | 0.000 | 18:31 |                      |
| Diagnostics Check | OK    | 18:31 |                      |

---

Data Download Date/Time: 04/29/2016 17:51

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SPENCE,KEITH,D  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: SPENCE,KEITH  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 03/24/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1686772  
DRY GAS EXPIRY DATE: 06/02/2017  
VIOLATION CODE: DUI  
ARREST TIME: 17:30  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 17:40

SUBJECT'S LAST NAME: PARRISH

DATE/TIME OF TEST: 04/24/2016 18:15:46

---

|                   |           |       |                |
|-------------------|-----------|-------|----------------|
| Diagnostics Check | OK        | 18:00 |                |
| Air Blank         | 0.000     | 18:00 |                |
| Control Test      | 0.082     | 18:01 |                |
| Air Blank         | 0.000     | 18:01 |                |
| Subject Sample    | # 1 0.341 | 18:02 |                |
| Breath Volume     | 2.925     | 18:02 |                |
| Air Blank         | 0.000     | 18:03 |                |
| Air Blank         | 0.000     | 18:04 |                |
| Subject Sample    | # 2 0.337 | 18:08 | Volume Not Met |
| Breath Volume     | 0.074     | 18:08 |                |
| Air Blank         | 0.000     | 18:08 |                |
| Air Blank         | 0.000     | 18:10 |                |
| Subject Sample    | # 3 0.333 | 18:13 | Volume Not Met |
| Breath Volume     | 0.062     | 18:13 |                |
| Air Blank         | 0.000     | 18:14 |                |
| Control Test      | 0.080     | 18:15 |                |
| Air Blank         | 0.000     | 18:15 |                |
| Diagnostics Check | OK        | 18:15 |                |

---

Data Download Date/Time: 04/29/2016 17:51

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HUGHES,ALFRED,L  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: HUGHES,ALFRED  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 04/29/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1686772  
DRY GAS EXPIRY DATE: 06/02/2017  
VIOLATION CODE: DUI  
ARREST TIME: 01:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:17

SUBJECT'S LAST NAME: JOUANNIC

DATE/TIME OF TEST: 04/30/2016 04:48:31

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:42 |
| Air Blank          | 0.000 | 04:42 |
| Control Test       | 0.082 | 04:43 |
| Air Blank          | 0.000 | 04:43 |
| Subject Sample # 1 | 0.209 | 04:44 |
| Breath Volume      | 2.507 | 04:44 |
| Air Blank          | 0.000 | 04:44 |
| Air Blank          | 0.000 | 04:46 |
| Subject Sample # 2 | 0.204 | 04:46 |
| Breath Volume      | 2.042 | 04:46 |
| Air Blank          | 0.000 | 04:47 |
| Control Test       | 0.081 | 04:47 |
| Air Blank          | 0.000 | 04:48 |
| Diagnostics Check  | OK    | 04:48 |

---

Data Download Date/Time: 05/24/2016 18:15

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000780

INSTRUMENT REGISTERED TO: SUWANNEE CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARRS,RUSSELL,B  
BREATH TEST OPERATOR AGENCY: Suwannee County SO  
ARRESTING OFFICER: CLARK,ZAC  
ARRESTING OFFICER AGENCY: Suwannee County SO  
LAST AGENCY INSPECTION: 04/01/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06415080A1  
DRY GAS EXPIRY DATE: 04/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 23:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: MACDONALD

DATE/TIME OF TEST: 04/03/2016 00:29:54

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:21 |
| Air Blank          | 0.000 | 00:22 |
| Control Test       | 0.083 | 00:22 |
| Air Blank          | 0.000 | 00:23 |
| Subject Sample # 1 | 0.154 | 00:25 |
| Breath Volume      | 1.902 | 00:25 |
| Air Blank          | 0.000 | 00:26 |
| Air Blank          | 0.000 | 00:27 |
| Subject Sample # 2 | 0.165 | 00:28 |
| Breath Volume      | 2.195 | 00:28 |
| Air Blank          | 0.000 | 00:28 |
| Control Test       | 0.079 | 00:29 |
| Air Blank          | 0.000 | 00:29 |
| Diagnostics Check  | OK    | 00:29 |

---

Data Download Date/Time: 05/31/2016 22:33



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000780

INSTRUMENT REGISTERED TO: SUWANNEE CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARRS,RUSSELL,B  
BREATH TEST OPERATOR AGENCY: Suwannee County SO  
ARRESTING OFFICER: WILLIAMSON,JACOB  
ARRESTING OFFICER AGENCY: Live Oak PD  
LAST AGENCY INSPECTION: 04/01/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06415080A1  
DRY GAS EXPIRY DATE: 04/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 04:04  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:14

SUBJECT'S LAST NAME: PITTMAN

DATE/TIME OF TEST: 04/03/2016 04:46:54

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:40 |
| Air Blank          | 0.000 | 04:40 |
| Control Test       | 0.083 | 04:40 |
| Air Blank          | 0.000 | 04:41 |
| Subject Sample # 1 | 0.143 | 04:42 |
| Breath Volume      | 1.832 | 04:42 |
| Air Blank          | 0.000 | 04:43 |
| Air Blank          | 0.000 | 04:44 |
| Subject Sample # 2 | 0.140 | 04:45 |
| Breath Volume      | 1.835 | 04:45 |
| Air Blank          | 0.000 | 04:45 |
| Control Test       | 0.081 | 04:46 |
| Air Blank          | 0.000 | 04:46 |
| Diagnostics Check  | OK    | 04:46 |

---

Data Download Date/Time: 05/31/2016 22:33

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000780

INSTRUMENT REGISTERED TO: SUWANNEE CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CLARK,J,T  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: CLARK,J  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 04/01/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06415080A1  
DRY GAS EXPIRY DATE: 04/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 20:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: BAKER

DATE/TIME OF TEST: 04/06/2016 21:34:10

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 21:27 |
| Air Blank          | 0.000 | 21:28 |
| Control Test       | 0.082 | 21:28 |
| Air Blank          | 0.000 | 21:29 |
| Subject Sample # 1 | 0.140 | 21:29 |
| Breath Volume      | 3.359 | 21:29 |
| Air Blank          | 0.000 | 21:30 |
| Air Blank          | 0.000 | 21:32 |
| Subject Sample # 2 | 0.130 | 21:32 |
| Breath Volume      | 1.765 | 21:32 |
| Air Blank          | 0.000 | 21:33 |
| Control Test       | 0.082 | 21:33 |
| Air Blank          | 0.000 | 21:34 |
| Diagnostics Check  | OK    | 21:34 |

---

Data Download Date/Time: 05/31/2016 22:33

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000780

INSTRUMENT REGISTERED TO: SUWANNEE CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ZIMMERMANN,JOHN,C  
BREATH TEST OPERATOR AGENCY: Suwannee County SO  
ARRESTING OFFICER: BROOKS,JAKE  
ARRESTING OFFICER AGENCY: Suwannee County SO  
LAST AGENCY INSPECTION: 04/01/2016

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 06415080A1  
DRY GAS EXPIRY DATE: 04/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 15:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 15:17

SUBJECT'S LAST NAME: YUHAS

DATE/TIME OF TEST: 04/08/2016 15:48:45

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 15:45 |                      |
| Air Blank         | 0.000 | 15:46 |                      |
| Control Test      | 0.082 | 15:46 |                      |
| Air Blank         | 0.000 | 15:47 |                      |
| Subject Sample    | # 1   | 15:47 | Subject Test Refused |
| Breath Volume     | 0.000 | 15:47 |                      |
| Air Blank         | 0.000 | 15:47 |                      |
| Control Test      | 0.082 | 15:48 |                      |
| Air Blank         | 0.000 | 15:48 |                      |
| Diagnostics Check | OK    | 15:48 |                      |

---

Data Download Date/Time: 05/31/2016 22:33

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000780

INSTRUMENT REGISTERED TO: SUWANNEE CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WADFORD,CHARLES,T  
BREATH TEST OPERATOR AGENCY: Suwannee County SO  
ARRESTING OFFICER: WADFORD,CHARLES  
ARRESTING OFFICER AGENCY: Suwannee County SO  
LAST AGENCY INSPECTION: 04/01/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06415080A1  
DRY GAS EXPIRY DATE: 04/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 02:30  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: CONWAY

DATE/TIME OF TEST: 04/14/2016 03:16:32

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:10 |
| Air Blank          | 0.000 | 03:10 |
| Control Test       | 0.082 | 03:11 |
| Air Blank          | 0.000 | 03:11 |
| Subject Sample # 1 | 0.207 | 03:12 |
| Breath Volume      | 1.882 | 03:12 |
| Air Blank          | 0.000 | 03:12 |
| Air Blank          | 0.000 | 03:14 |
| Subject Sample # 2 | 0.202 | 03:14 |
| Breath Volume      | 2.046 | 03:14 |
| Air Blank          | 0.000 | 03:15 |
| Control Test       | 0.083 | 03:15 |
| Air Blank          | 0.000 | 03:16 |
| Diagnostics Check  | OK    | 03:16 |

---

Data Download Date/Time: 05/31/2016 22:33

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000780

INSTRUMENT REGISTERED TO: SUWANNEE CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KELLY,DAVID,W  
BREATH TEST OPERATOR AGENCY: Suwannee County SO  
ARRESTING OFFICER: CARTER,DANIEL  
ARRESTING OFFICER AGENCY: Suwannee County SO  
LAST AGENCY INSPECTION: 04/01/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06415080A1  
DRY GAS EXPIRY DATE: 04/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 03:30  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:06

SUBJECT'S LAST NAME: EAGLE

DATE/TIME OF TEST: 04/17/2016 04:37:58

---

|                   |           |       |            |
|-------------------|-----------|-------|------------|
| Diagnostics Check | OK        | 04:32 |            |
| Air Blank         | 0.000     | 04:33 |            |
| Control Test      | 0.082     | 04:33 |            |
| Air Blank         | 0.000     | 04:33 |            |
| Subject Sample    | # 1 0.172 | 04:36 |            |
| Breath Volume     | 2.734     | 04:36 |            |
| Air Blank         |           | 04:37 | Purge Fail |
| Air Blank         | 0.000     | 04:37 |            |

---

Data Download Date/Time: 05/31/2016 22:33

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000780

INSTRUMENT REGISTERED TO: SUWANNEE CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CLARK,J,T  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: CLARK,J  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 04/01/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06415080A1  
DRY GAS EXPIRY DATE: 04/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 19:35  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 19:40

SUBJECT'S LAST NAME: BLAKE

DATE/TIME OF TEST: 04/20/2016 20:15:36

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 20:09 |
| Air Blank          | 0.000 | 20:09 |
| Control Test       | 0.083 | 20:09 |
| Air Blank          | 0.000 | 20:10 |
| Subject Sample # 1 | 0.135 | 20:10 |
| Breath Volume      | 1.457 | 20:10 |
| Air Blank          | 0.000 | 20:11 |
| Air Blank          | 0.000 | 20:13 |
| Subject Sample # 2 | 0.126 | 20:14 |
| Breath Volume      | 1.468 | 20:14 |
| Air Blank          | 0.000 | 20:14 |
| Control Test       | 0.083 | 20:15 |
| Air Blank          | 0.000 | 20:15 |
| Diagnostics Check  | OK    | 20:15 |

---

Data Download Date/Time: 05/31/2016 22:33

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000780

INSTRUMENT REGISTERED TO: SUWANNEE CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HENRY,MICHAEL,S  
BREATH TEST OPERATOR AGENCY: Suwannee County SO  
ARRESTING OFFICER: HENRY,MICHAEL  
ARRESTING OFFICER AGENCY: Suwannee County SO  
LAST AGENCY INSPECTION: 04/01/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06415080A1  
DRY GAS EXPIRY DATE: 04/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 21:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: REQUESENS ORTIZ

DATE/TIME OF TEST: 04/22/2016 22:30:57

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:24 |
| Air Blank          | 0.000 | 22:24 |
| Control Test       | 0.082 | 22:24 |
| Air Blank          | 0.000 | 22:25 |
| Subject Sample # 1 | 0.264 | 22:26 |
| Breath Volume      | 1.628 | 22:26 |
| Air Blank          | 0.000 | 22:26 |
| Air Blank          | 0.000 | 22:28 |
| Subject Sample # 2 | 0.254 | 22:29 |
| Breath Volume      | 1.292 | 22:29 |
| Air Blank          | 0.000 | 22:30 |
| Control Test       | 0.080 | 22:30 |
| Air Blank          | 0.000 | 22:30 |
| Diagnostics Check  | OK    | 22:30 |

---

Data Download Date/Time: 05/31/2016 22:34

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000780

INSTRUMENT REGISTERED TO: SUWANNEE CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HUGHES,ALFRED,L  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: HUGHES,ALFRED  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 04/01/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06415080A1  
DRY GAS EXPIRY DATE: 04/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 02:56  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:17

SUBJECT'S LAST NAME: WARD

DATE/TIME OF TEST: 04/27/2016 04:49:13

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:42 |
| Air Blank          | 0.000 | 04:43 |
| Control Test       | 0.083 | 04:43 |
| Air Blank          | 0.000 | 04:44 |
| Subject Sample # 1 | 0.112 | 04:44 |
| Breath Volume      | 2.613 | 04:44 |
| Air Blank          | 0.000 | 04:45 |
| Air Blank          | 0.000 | 04:47 |
| Subject Sample # 2 | 0.102 | 04:47 |
| Breath Volume      | 2.148 | 04:47 |
| Air Blank          | 0.000 | 04:48 |
| Control Test       | 0.081 | 04:48 |
| Air Blank          | 0.000 | 04:49 |
| Diagnostics Check  | OK    | 04:49 |

---

Data Download Date/Time: 05/31/2016 22:34



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000780

INSTRUMENT REGISTERED TO: SUWANNEE CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WADFORD,CHARLES,T  
BREATH TEST OPERATOR AGENCY: Suwannee County SO  
ARRESTING OFFICER: WADFORD,CHARLES  
ARRESTING OFFICER AGENCY: Suwannee County SO  
LAST AGENCY INSPECTION: 04/01/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06415080A1  
DRY GAS EXPIRY DATE: 04/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 20:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:45

SUBJECT'S LAST NAME: CRAPPS

DATE/TIME OF TEST: 04/30/2016 21:26:08

---

|                    |       |       |               |
|--------------------|-------|-------|---------------|
| Diagnostics Check  | OK    | 21:19 |               |
| Air Blank          | 0.000 | 21:20 |               |
| Control Test       | 0.082 | 21:20 |               |
| Air Blank          | 0.000 | 21:21 |               |
| Subject Sample # 1 | 0.411 | 21:21 |               |
| Breath Volume      | 3.140 | 21:21 |               |
| Air Blank          | 0.000 | 21:22 |               |
| Air Blank          | 0.000 | 21:24 |               |
| Subject Sample # 2 |       | 21:24 | Slope Not Met |
| Breath Volume      | 1.984 | 21:24 |               |
| Air Blank          | 0.000 | 21:25 |               |
| Control Test       | 0.081 | 21:25 |               |
| Air Blank          | 0.000 | 21:26 |               |
| Diagnostics Check  | OK    | 21:26 |               |

---

Data Download Date/Time: 05/31/2016 22:34

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000780

INSTRUMENT REGISTERED TO: SUWANNEE CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WADFORD,CHARLES,T  
BREATH TEST OPERATOR AGENCY: Suwannee County SO  
ARRESTING OFFICER: WADFORD,CHARLES  
ARRESTING OFFICER AGENCY: Suwannee County SO  
LAST AGENCY INSPECTION: 04/01/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06415080A1  
DRY GAS EXPIRY DATE: 04/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 20:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:45

SUBJECT'S LAST NAME: CRAPPS

DATE/TIME OF TEST: 04/30/2016 21:34:36

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 21:28 |
| Air Blank          | 0.000 | 21:28 |
| Control Test       | 0.082 | 21:29 |
| Air Blank          | 0.000 | 21:29 |
| Subject Sample # 1 | 0.398 | 21:30 |
| Breath Volume      | 3.875 | 21:30 |
| Air Blank          | 0.000 | 21:30 |
| Air Blank          | 0.000 | 21:32 |
| Subject Sample # 2 | 0.404 | 21:32 |
| Breath Volume      | 3.355 | 21:32 |
| Air Blank          | 0.000 | 21:33 |
| Control Test       | 0.082 | 21:33 |
| Air Blank          | 0.000 | 21:34 |
| Diagnostics Check  | OK    | 21:34 |

---

Data Download Date/Time: 05/31/2016 22:34

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000785

INSTRUMENT REGISTERED TO: HOLMES COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARSEY, WESLEY, H  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: HARSEY, WESLEY  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 03/05/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 20314080A4  
DRY GAS EXPIRY DATE: 08/05/2016  
VIOLATION CODE: DUI  
ARREST TIME: 00:45  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:11

SUBJECT'S LAST NAME: HOLLOWAY

DATE/TIME OF TEST: 04/20/2016 02:43:51

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:36 |
| Air Blank          | 0.000 | 02:36 |
| Control Test       | 0.079 | 02:37 |
| Air Blank          | 0.000 | 02:37 |
| Subject Sample # 1 | 0.153 | 02:39 |
| Breath Volume      | 1.210 | 02:39 |
| Air Blank          | 0.000 | 02:39 |
| Air Blank          | 0.000 | 02:41 |
| Subject Sample # 2 | 0.152 | 02:42 |
| Breath Volume      | 2.363 | 02:42 |
| Air Blank          | 0.000 | 02:42 |
| Control Test       | 0.078 | 02:43 |
| Air Blank          | 0.000 | 02:43 |
| Diagnostics Check  | OK    | 02:43 |

---

Data Download Date/Time: 07/03/2016 16:16

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000785

INSTRUMENT REGISTERED TO: HOLMES COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: VENNELL,JOHN,B  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: VENNELL,JOHN  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 03/05/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 20314080A4  
DRY GAS EXPIRY DATE: 08/05/2016  
VIOLATION CODE: DUI  
ARREST TIME: 02:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: PHILLIPS

DATE/TIME OF TEST: 04/24/2016 03:39:30

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 03:36 |                      |
| Air Blank         | 0.000 | 03:37 |                      |
| Control Test      | 0.080 | 03:37 |                      |
| Air Blank         | 0.000 | 03:37 |                      |
| Subject Sample    | # 1   | 03:38 | Subject Test Refused |
| Breath Volume     | 0.000 | 03:38 |                      |
| Air Blank         | 0.000 | 03:38 |                      |
| Control Test      | 0.079 | 03:38 |                      |
| Air Blank         | 0.000 | 03:39 |                      |
| Diagnostics Check | OK    | 03:39 |                      |

---

Data Download Date/Time: 07/03/2016 16:16

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NGUYEN,TAI,T  
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD  
ARRESTING OFFICER: NGUYEN,TAI  
ARRESTING OFFICER AGENCY: Gulf Breeze PD  
LAST AGENCY INSPECTION: 03/15/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18015080A2  
DRY GAS EXPIRY DATE: 08/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 01:48  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: IMHOLT

DATE/TIME OF TEST: 04/10/2016 02:32:15

---

|                    |       |       |            |
|--------------------|-------|-------|------------|
| Diagnostics Check  | OK    | 02:28 |            |
| Air Blank          | 0.000 | 02:28 |            |
| Control Test       | 0.078 | 02:28 |            |
| Air Blank          | 0.000 | 02:29 |            |
| Subject Sample # 1 | 0.188 | 02:30 |            |
| Breath Volume      | 2.046 | 02:30 |            |
| Air Blank          |       | 02:31 | Purge Fail |
| Air Blank          | 0.000 | 02:32 |            |

---

Data Download Date/Time: 04/15/2016 10:31

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NGUYEN,TAI,T  
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD  
ARRESTING OFFICER: NGUYEN,TAI  
ARRESTING OFFICER AGENCY: Gulf Breeze PD  
LAST AGENCY INSPECTION: 03/15/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18015080A2  
DRY GAS EXPIRY DATE: 08/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 01:48  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: IMHOLT

DATE/TIME OF TEST: 04/10/2016 02:38:48

---

|                   |           |       |            |
|-------------------|-----------|-------|------------|
| Diagnostics Check | OK        | 02:34 |            |
| Air Blank         | 0.000     | 02:34 |            |
| Control Test      | 0.078     | 02:35 |            |
| Air Blank         | 0.000     | 02:35 |            |
| Subject Sample    | # 1 0.176 | 02:37 |            |
| Breath Volume     | 1.410     | 02:37 |            |
| Air Blank         |           | 02:37 | Purge Fail |
| Air Blank         | 0.000     | 02:38 |            |

---

Data Download Date/Time: 04/15/2016 10:31

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NGUYEN,TAI,T  
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD  
ARRESTING OFFICER: NGUYEN,TAI  
ARRESTING OFFICER AGENCY: Gulf Breeze PD  
LAST AGENCY INSPECTION: 03/15/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18015080A2  
DRY GAS EXPIRY DATE: 08/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 01:48  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: IMHOLT

DATE/TIME OF TEST: 04/10/2016 02:54:58

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:48 |
| Air Blank          | 0.000 | 02:49 |
| Control Test       | 0.077 | 02:49 |
| Air Blank          | 0.000 | 02:50 |
| Subject Sample # 1 | 0.189 | 02:50 |
| Breath Volume      | 1.167 | 02:50 |
| Air Blank          | 0.000 | 02:51 |
| Air Blank          | 0.000 | 02:52 |
| Subject Sample # 2 | 0.180 | 02:53 |
| Breath Volume      | 1.621 | 02:53 |
| Air Blank          | 0.000 | 02:54 |
| Control Test       | 0.078 | 02:54 |
| Air Blank          | 0.000 | 02:54 |
| Diagnostics Check  | OK    | 02:54 |

---

Data Download Date/Time: 04/15/2016 10:31

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WHITE,TRAVIS,D  
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD  
ARRESTING OFFICER: WHITE,TRAVIS  
ARRESTING OFFICER AGENCY: Gulf Breeze PD  
LAST AGENCY INSPECTION: 04/15/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18015080A2  
DRY GAS EXPIRY DATE: 08/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 18:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:36

SUBJECT'S LAST NAME: LALAS

DATE/TIME OF TEST: 04/17/2016 19:02:46

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 18:56 |
| Air Blank          | 0.000 | 18:56 |
| Control Test       | 0.079 | 18:57 |
| Air Blank          | 0.000 | 18:57 |
| Subject Sample # 1 | 0.190 | 18:58 |
| Breath Volume      | 3.546 | 18:58 |
| Air Blank          | 0.000 | 18:58 |
| Air Blank          | 0.000 | 19:00 |
| Subject Sample # 2 | 0.184 | 19:01 |
| Breath Volume      | 3.058 | 19:01 |
| Air Blank          | 0.000 | 19:01 |
| Control Test       | 0.079 | 19:02 |
| Air Blank          | 0.000 | 19:02 |
| Diagnostics Check  | OK    | 19:02 |

---

Data Download Date/Time: 05/17/2016 03:02



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TAVEIRNE,ROBERT,J  
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD  
ARRESTING OFFICER: KIDD,J  
ARRESTING OFFICER AGENCY: Gulf Breeze PD  
LAST AGENCY INSPECTION: 04/15/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18015080A2  
DRY GAS EXPIRY DATE: 08/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 21:56  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: HARMON

DATE/TIME OF TEST: 04/24/2016 22:57:23

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 22:54 |                      |
| Air Blank         | 0.000 | 22:55 |                      |
| Control Test      | 0.077 | 22:55 |                      |
| Air Blank         | 0.000 | 22:55 |                      |
| Subject Sample    | # 1   | 22:56 | Subject Test Refused |
| Breath Volume     | 0.000 | 22:56 |                      |
| Air Blank         | 0.000 | 22:56 |                      |
| Control Test      | 0.078 | 22:56 |                      |
| Air Blank         | 0.000 | 22:57 |                      |
| Diagnostics Check | OK    | 22:57 |                      |

---

Data Download Date/Time: 05/17/2016 03:02

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SCAPIN,BRANDON,L  
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD  
ARRESTING OFFICER: SCAPIN,BRANDON  
ARRESTING OFFICER AGENCY: Gulf Breeze PD  
LAST AGENCY INSPECTION: 04/15/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18015080A2  
DRY GAS EXPIRY DATE: 08/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 19:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:58

SUBJECT'S LAST NAME: REED

DATE/TIME OF TEST: 04/25/2016 20:28:36

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 20:18 |
| Air Blank          | 0.000 | 20:18 |
| Control Test       | 0.079 | 20:19 |
| Air Blank          | 0.000 | 20:19 |
| Subject Sample # 1 | 0.123 | 20:20 |
| Breath Volume      | 4.281 | 20:20 |
| Air Blank          | 0.000 | 20:21 |
| Air Blank          | 0.000 | 20:23 |
| Subject Sample # 2 | 0.098 | 20:23 |
| Breath Volume      | 1.769 | 20:23 |
| Air Blank          | 0.000 | 20:24 |
| Air Blank          | 0.000 | 20:26 |
| Subject Sample # 3 | 0.116 | 20:27 |
| Breath Volume      | 1.582 | 20:27 |
| Air Blank          | 0.000 | 20:27 |
| Control Test       | 0.080 | 20:28 |
| Air Blank          | 0.000 | 20:28 |
| Diagnostics Check  | OK    | 20:28 |

---

Data Download Date/Time: 05/17/2016 03:02

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SCAPIN,BRANDON,L  
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD  
ARRESTING OFFICER: SCAPIN,BRANDON  
ARRESTING OFFICER AGENCY: Gulf Breeze PD  
LAST AGENCY INSPECTION: 04/15/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18015080A2  
DRY GAS EXPIRY DATE: 08/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 23:44  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: MARCANIO

DATE/TIME OF TEST: 04/30/2016 00:37:42

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:31 |
| Air Blank          | 0.000 | 00:31 |
| Control Test       | 0.079 | 00:31 |
| Air Blank          | 0.000 | 00:32 |
| Subject Sample # 1 | 0.186 | 00:33 |
| Breath Volume      | 2.531 | 00:33 |
| Air Blank          | 0.000 | 00:33 |
| Air Blank          | 0.000 | 00:35 |
| Subject Sample # 2 | 0.185 | 00:36 |
| Breath Volume      | 2.171 | 00:36 |
| Air Blank          | 0.000 | 00:36 |
| Control Test       | 0.077 | 00:37 |
| Air Blank          | 0.000 | 00:37 |
| Diagnostics Check  | OK    | 00:37 |

---

Data Download Date/Time: 05/17/2016 03:02

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000790

INSTRUMENT REGISTERED TO: VERO BEACH P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KNUTSON,NICHOLAS,A  
BREATH TEST OPERATOR AGENCY: Vero Beach PD  
ARRESTING OFFICER: KNUTSON,NICHOLAS  
ARRESTING OFFICER AGENCY: Vero Beach PD  
LAST AGENCY INSPECTION: 03/26/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 16014080A2  
DRY GAS EXPIRY DATE: 07/05/2016  
VIOLATION CODE: DUI  
ARREST TIME: 01:19  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:43

SUBJECT'S LAST NAME: WICKER

DATE/TIME OF TEST: 04/02/2016 02:07:21

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 02:04 |                      |
| Air Blank         | 0.000 | 02:04 |                      |
| Control Test      | 0.080 | 02:05 |                      |
| Air Blank         | 0.000 | 02:05 |                      |
| Subject Sample    | # 1   | 02:06 | Subject Test Refused |
| Breath Volume     | 0.000 | 02:06 |                      |
| Air Blank         | 0.000 | 02:06 |                      |
| Control Test      | 0.079 | 02:06 |                      |
| Air Blank         | 0.000 | 02:07 |                      |
| Diagnostics Check | OK    | 02:07 |                      |

---

Data Download Date/Time: 04/29/2016 04:57

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000790

INSTRUMENT REGISTERED TO: VERO BEACH P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DUDA,IGA,  
BREATH TEST OPERATOR AGENCY: Vero Beach PD  
ARRESTING OFFICER: HARRIS,JOSHUA  
ARRESTING OFFICER AGENCY: Vero Beach PD  
LAST AGENCY INSPECTION: 03/26/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 16014080A2  
DRY GAS EXPIRY DATE: 07/05/2016  
VIOLATION CODE: DUI  
ARREST TIME: 02:33  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:58

SUBJECT'S LAST NAME: DUDA

DATE/TIME OF TEST: 04/02/2016 03:26:10

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:18 |
| Air Blank          | 0.000 | 03:19 |
| Control Test       | 0.079 | 03:19 |
| Air Blank          | 0.000 | 03:20 |
| Subject Sample # 1 | 0.172 | 03:21 |
| Breath Volume      | 1.613 | 03:21 |
| Air Blank          | 0.000 | 03:21 |
| Air Blank          | 0.000 | 03:23 |
| Subject Sample # 2 | 0.163 | 03:24 |
| Breath Volume      | 2.257 | 03:24 |
| Air Blank          | 0.000 | 03:25 |
| Control Test       | 0.078 | 03:25 |
| Air Blank          | 0.000 | 03:26 |
| Diagnostics Check  | OK    | 03:26 |

---

Data Download Date/Time: 04/29/2016 04:57

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000790

INSTRUMENT REGISTERED TO: VERO BEACH P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: EGGERS,RYAN,K  
BREATH TEST OPERATOR AGENCY: Vero Beach PD  
ARRESTING OFFICER: EGGERS,RYAN  
ARRESTING OFFICER AGENCY: Vero Beach PD  
LAST AGENCY INSPECTION: 03/26/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 16014080A2  
DRY GAS EXPIRY DATE: 07/05/2016  
VIOLATION CODE: DUI  
ARREST TIME: 20:58  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 20:59

SUBJECT'S LAST NAME: PRATT

DATE/TIME OF TEST: 04/07/2016 21:31:26

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 21:23 |
| Air Blank          | 0.000 | 21:23 |
| Control Test       | 0.080 | 21:24 |
| Air Blank          | 0.000 | 21:24 |
| Subject Sample # 1 | 0.284 | 21:26 |
| Breath Volume      | 1.714 | 21:26 |
| Air Blank          | 0.000 | 21:26 |
| Air Blank          | 0.000 | 21:28 |
| Subject Sample # 2 | 0.264 | 21:29 |
| Breath Volume      | 1.179 | 21:29 |
| Air Blank          | 0.000 | 21:30 |
| Control Test       | 0.079 | 21:30 |
| Air Blank          | 0.000 | 21:31 |
| Diagnostics Check  | OK    | 21:31 |

---

Data Download Date/Time: 04/29/2016 04:57

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000790

INSTRUMENT REGISTERED TO: VERO BEACH P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LOHMAN,JENNIFER,L  
BREATH TEST OPERATOR AGENCY: Vero Beach PD  
ARRESTING OFFICER: LOHMAN,JENNIFER  
ARRESTING OFFICER AGENCY: Vero Beach PD  
LAST AGENCY INSPECTION: 03/26/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 16014080A2  
DRY GAS EXPIRY DATE: 07/05/2016  
VIOLATION CODE: DUI  
ARREST TIME: 18:52  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:17

SUBJECT'S LAST NAME: ARANDA

DATE/TIME OF TEST: 04/10/2016 19:42:07

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 19:39 |                      |
| Air Blank         | 0.000 | 19:39 |                      |
| Control Test      | 0.080 | 19:40 |                      |
| Air Blank         | 0.000 | 19:40 |                      |
| Subject Sample    | # 1   | 19:40 | Subject Test Refused |
| Breath Volume     | 0.000 | 19:40 |                      |
| Air Blank         | 0.000 | 19:41 |                      |
| Control Test      | 0.079 | 19:41 |                      |
| Air Blank         | 0.000 | 19:42 |                      |
| Diagnostics Check | OK    | 19:42 |                      |

---

Data Download Date/Time: 04/29/2016 04:57

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000790

INSTRUMENT REGISTERED TO: VERO BEACH P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARRIS,JOSHUA,D  
BREATH TEST OPERATOR AGENCY: Vero Beach PD  
ARRESTING OFFICER: HARRIS,JOSHUA  
ARRESTING OFFICER AGENCY: Vero Beach PD  
LAST AGENCY INSPECTION: 03/26/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 16014080A2  
DRY GAS EXPIRY DATE: 07/05/2016  
VIOLATION CODE: DUI  
ARREST TIME: 23:08  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:33

SUBJECT'S LAST NAME: ROTH

DATE/TIME OF TEST: 04/11/2016 23:59:39

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:53 |
| Air Blank          | 0.000 | 23:53 |
| Control Test       | 0.080 | 23:54 |
| Air Blank          | 0.000 | 23:54 |
| Subject Sample # 1 | 0.000 | 23:55 |
| Breath Volume      | 3.359 | 23:55 |
| Air Blank          | 0.000 | 23:55 |
| Air Blank          | 0.000 | 23:57 |
| Subject Sample # 2 | 0.000 | 23:58 |
| Breath Volume      | 4.542 | 23:58 |
| Air Blank          | 0.000 | 23:58 |
| Control Test       | 0.078 | 23:59 |
| Air Blank          | 0.000 | 23:59 |
| Diagnostics Check  | OK    | 23:59 |

---

Data Download Date/Time: 04/29/2016 04:57



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000790

INSTRUMENT REGISTERED TO: VERO BEACH P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GINSBURG,DOUG,J  
BREATH TEST OPERATOR AGENCY: Vero Beach PD  
ARRESTING OFFICER: GINSBURG,DOUG  
ARRESTING OFFICER AGENCY: Vero Beach PD  
LAST AGENCY INSPECTION: 03/26/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 16014080A2  
DRY GAS EXPIRY DATE: 07/05/2016  
VIOLATION CODE: DUI  
ARREST TIME: 23:19  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:19

SUBJECT'S LAST NAME: SHEETS

DATE/TIME OF TEST: 04/29/2016 00:09:47

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 00:07 |                      |
| Air Blank         | 0.000 | 00:07 |                      |
| Control Test      | 0.080 | 00:07 |                      |
| Air Blank         | 0.000 | 00:08 |                      |
| Subject Sample    | # 1   | 00:08 | Subject Test Refused |
| Breath Volume     | 0.000 | 00:08 |                      |
| Air Blank         | 0.000 | 00:08 |                      |
| Control Test      | 0.079 | 00:09 |                      |
| Air Blank         | 0.000 | 00:09 |                      |
| Diagnostics Check | OK    | 00:09 |                      |

---

Data Download Date/Time: 04/29/2016 04:57

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MAZANOSKI, MITCHELL, S  
BREATH TEST OPERATOR AGENCY: St. Lucie County SO  
ARRESTING OFFICER: MAZANOSKI, MITCHELL  
ARRESTING OFFICER AGENCY: St. Lucie County SO  
LAST AGENCY INSPECTION: 03/16/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26614080A7  
DRY GAS EXPIRY DATE: 10/05/2016  
VIOLATION CODE: DUI  
ARREST TIME: 02:06  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:06

SUBJECT'S LAST NAME: HART

DATE/TIME OF TEST: 04/01/2016 02:48:35

---

|                   |           |       |            |
|-------------------|-----------|-------|------------|
| Diagnostics Check | OK        | 02:45 |            |
| Air Blank         | 0.000     | 02:45 |            |
| Control Test      | 0.080     | 02:45 |            |
| Air Blank         | 0.000     | 02:46 |            |
| Subject Sample    | # 1 0.202 | 02:46 |            |
| Breath Volume     | 2.570     | 02:46 |            |
| Air Blank         |           | 02:47 | RFI Detect |
| Air Blank         | 0.000     | 02:47 |            |
| Control Test      | 0.080     | 02:47 |            |
| Air Blank         | 0.000     | 02:48 |            |
| Diagnostics Check | OK        | 02:48 |            |

---

Data Download Date/Time: 04/28/2016 13:22

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MAZANOSKI, MITCHELL, S  
BREATH TEST OPERATOR AGENCY: St. Lucie County SO  
ARRESTING OFFICER: MAZANOSKI, MITCHELL  
ARRESTING OFFICER AGENCY: St. Lucie County SO  
LAST AGENCY INSPECTION: 03/16/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26614080A7  
DRY GAS EXPIRY DATE: 10/05/2016  
VIOLATION CODE: DUI  
ARREST TIME: 02:06  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:06

SUBJECT'S LAST NAME: HART

DATE/TIME OF TEST: 04/01/2016 02:57:28

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:50 |
| Air Blank          | 0.000 | 02:51 |
| Control Test       | 0.079 | 02:51 |
| Air Blank          | 0.000 | 02:51 |
| Subject Sample # 1 | 0.202 | 02:52 |
| Breath Volume      | 2.222 | 02:52 |
| Air Blank          | 0.000 | 02:53 |
| Air Blank          | 0.000 | 02:55 |
| Subject Sample # 2 | 0.199 | 02:55 |
| Breath Volume      | 2.152 | 02:55 |
| Air Blank          | 0.000 | 02:56 |
| Control Test       | 0.078 | 02:56 |
| Air Blank          | 0.000 | 02:57 |
| Diagnostics Check  | OK    | 02:57 |

---

Data Download Date/Time: 04/28/2016 13:22

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HOGAN,TODD,A  
BREATH TEST OPERATOR AGENCY: St. Lucie County SO  
ARRESTING OFFICER: HOGAN,TODD  
ARRESTING OFFICER AGENCY: St. Lucie County SO  
LAST AGENCY INSPECTION: 03/16/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26614080A7  
DRY GAS EXPIRY DATE: 10/05/2016  
VIOLATION CODE: DUI  
ARREST TIME: 18:59  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 19:36

SUBJECT'S LAST NAME: KLINGLER

DATE/TIME OF TEST: 04/07/2016 20:09:17

---

|                   |           |       |                 |
|-------------------|-----------|-------|-----------------|
| Diagnostics Check | OK        | 19:57 |                 |
| Air Blank         | 0.000     | 19:57 |                 |
| Control Test      | 0.079     | 19:57 |                 |
| Air Blank         | 0.000     | 19:58 |                 |
| Subject Sample    | # 1 0.091 | 20:01 | Slope Not Level |
| Breath Volume     | 1.925     | 20:01 |                 |
| Air Blank         | 0.000     | 20:02 |                 |
| Air Blank         | 0.000     | 20:04 |                 |
| Subject Sample    | # 2 0.093 | 20:04 |                 |
| Breath Volume     | 1.203     | 20:04 |                 |
| Air Blank         | 0.000     | 20:05 |                 |
| Air Blank         | 0.000     | 20:07 |                 |
| Subject Sample    | # 3 0.090 | 20:07 |                 |
| Breath Volume     | 1.835     | 20:07 |                 |
| Air Blank         | 0.000     | 20:08 |                 |
| Control Test      | 0.080     | 20:08 |                 |
| Air Blank         | 0.000     | 20:09 |                 |
| Diagnostics Check | OK        | 20:09 |                 |

---

Data Download Date/Time: 04/28/2016 13:22

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FELIX,GEORGE,D  
BREATH TEST OPERATOR AGENCY: St. Lucie County SO  
ARRESTING OFFICER: FELIX,GEORGE  
ARRESTING OFFICER AGENCY: St. Lucie County SO  
LAST AGENCY INSPECTION: 03/16/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26614080A7  
DRY GAS EXPIRY DATE: 10/05/2016  
VIOLATION CODE: DUI  
ARREST TIME: 00:07  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: SUTTON

DATE/TIME OF TEST: 04/11/2016 01:06:31

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:00 |
| Air Blank          | 0.000 | 01:00 |
| Control Test       | 0.081 | 01:00 |
| Air Blank          | 0.000 | 01:01 |
| Subject Sample # 1 | 0.153 | 01:02 |
| Breath Volume      | 1.320 | 01:02 |
| Air Blank          | 0.000 | 01:02 |
| Air Blank          | 0.000 | 01:04 |
| Subject Sample # 2 | 0.156 | 01:05 |
| Breath Volume      | 1.531 | 01:05 |
| Air Blank          | 0.000 | 01:05 |
| Control Test       | 0.079 | 01:05 |
| Air Blank          | 0.000 | 01:06 |
| Diagnostics Check  | OK    | 01:06 |

---

Data Download Date/Time: 04/28/2016 13:22

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FELIX,GEORGE,D  
BREATH TEST OPERATOR AGENCY: St. Lucie County SO  
ARRESTING OFFICER: FELIX,GEORGE  
ARRESTING OFFICER AGENCY: St. Lucie County SO  
LAST AGENCY INSPECTION: 03/16/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26614080A7  
DRY GAS EXPIRY DATE: 10/05/2016  
VIOLATION CODE: DUI  
ARREST TIME: 01:43  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: NASCIMENTO

DATE/TIME OF TEST: 04/15/2016 02:31:58

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:25 |
| Air Blank          | 0.000 | 02:25 |
| Control Test       | 0.080 | 02:26 |
| Air Blank          | 0.000 | 02:26 |
| Subject Sample # 1 | 0.181 | 02:27 |
| Breath Volume      | 1.781 | 02:27 |
| Air Blank          | 0.000 | 02:28 |
| Air Blank          | 0.000 | 02:29 |
| Subject Sample # 2 | 0.180 | 02:30 |
| Breath Volume      | 1.878 | 02:30 |
| Air Blank          | 0.000 | 02:31 |
| Control Test       | 0.079 | 02:31 |
| Air Blank          | 0.000 | 02:31 |
| Diagnostics Check  | OK    | 02:31 |

---

Data Download Date/Time: 04/28/2016 13:22

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HOGAN,TODD,A  
BREATH TEST OPERATOR AGENCY: St. Lucie County SO  
ARRESTING OFFICER: HOGAN,TODD  
ARRESTING OFFICER AGENCY: St. Lucie County SO  
LAST AGENCY INSPECTION: 03/16/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26614080A7  
DRY GAS EXPIRY DATE: 10/05/2016  
VIOLATION CODE: DUI  
ARREST TIME: 01:12  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:38

SUBJECT'S LAST NAME: CUADRA

DATE/TIME OF TEST: 04/16/2016 02:06:32

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:59 |
| Air Blank          | 0.000 | 01:59 |
| Control Test       | 0.079 | 01:59 |
| Air Blank          | 0.000 | 02:00 |
| Subject Sample # 1 | 0.229 | 02:00 |
| Breath Volume      | 1.617 | 02:00 |
| Air Blank          | 0.000 | 02:01 |
| Air Blank          | 0.000 | 02:03 |
| Subject Sample # 2 | 0.226 | 02:04 |
| Breath Volume      | 1.664 | 02:04 |
| Air Blank          | 0.000 | 02:05 |
| Control Test       | 0.080 | 02:05 |
| Air Blank          | 0.000 | 02:06 |
| Diagnostics Check  | OK    | 02:06 |

---

Data Download Date/Time: 04/28/2016 13:22

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FELIX,GEORGE,D  
BREATH TEST OPERATOR AGENCY: St. Lucie County SO  
ARRESTING OFFICER: FELIX,GEORGE  
ARRESTING OFFICER AGENCY: St. Lucie County SO  
LAST AGENCY INSPECTION: 03/16/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26614080A7  
DRY GAS EXPIRY DATE: 10/05/2016  
VIOLATION CODE: DUI  
ARREST TIME: 01:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: BOVA

DATE/TIME OF TEST: 04/16/2016 03:07:22

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:00 |
| Air Blank          | 0.000 | 03:01 |
| Control Test       | 0.080 | 03:01 |
| Air Blank          | 0.000 | 03:01 |
| Subject Sample # 1 | 0.151 | 03:02 |
| Breath Volume      | 4.347 | 03:02 |
| Air Blank          | 0.000 | 03:03 |
| Air Blank          | 0.000 | 03:05 |
| Subject Sample # 2 | 0.146 | 03:05 |
| Breath Volume      | 2.601 | 03:05 |
| Air Blank          | 0.000 | 03:06 |
| Control Test       | 0.078 | 03:06 |
| Air Blank          | 0.000 | 03:07 |
| Diagnostics Check  | OK    | 03:07 |

---

Data Download Date/Time: 04/28/2016 13:22



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CAGLIONI,DAVID,A  
BREATH TEST OPERATOR AGENCY: St. Lucie County SO  
ARRESTING OFFICER: CAGLIONI,DAVID  
ARRESTING OFFICER AGENCY: St. Lucie County SO  
LAST AGENCY INSPECTION: 03/16/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26614080A7  
DRY GAS EXPIRY DATE: 10/05/2016  
VIOLATION CODE: DUI  
ARREST TIME: 13:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 14:01

SUBJECT'S LAST NAME: DESCHAMPS

DATE/TIME OF TEST: 04/16/2016 14:27:54

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 14:21 |
| Air Blank          | 0.000 | 14:22 |
| Control Test       | 0.079 | 14:22 |
| Air Blank          | 0.000 | 14:23 |
| Subject Sample # 1 | 0.216 | 14:23 |
| Breath Volume      | 1.566 | 14:23 |
| Air Blank          | 0.000 | 14:24 |
| Air Blank          | 0.000 | 14:25 |
| Subject Sample # 2 | 0.213 | 14:26 |
| Breath Volume      | 1.546 | 14:26 |
| Air Blank          | 0.000 | 14:27 |
| Control Test       | 0.079 | 14:27 |
| Air Blank          | 0.000 | 14:27 |
| Diagnostics Check  | OK    | 14:27 |

---

Data Download Date/Time: 04/28/2016 13:22

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HOGAN,TODD,A  
BREATH TEST OPERATOR AGENCY: St. Lucie County SO  
ARRESTING OFFICER: HOGAN,TODD  
ARRESTING OFFICER AGENCY: St. Lucie County SO  
LAST AGENCY INSPECTION: 03/16/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26614080A7  
DRY GAS EXPIRY DATE: 10/05/2016  
VIOLATION CODE: DUI  
ARREST TIME: 23:14  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:38

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 04/17/2016 00:07:44

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:01 |
| Air Blank          | 0.000 | 00:01 |
| Control Test       | 0.080 | 00:02 |
| Air Blank          | 0.000 | 00:02 |
| Subject Sample # 1 | 0.205 | 00:03 |
| Breath Volume      | 1.707 | 00:03 |
| Air Blank          | 0.000 | 00:03 |
| Air Blank          | 0.000 | 00:05 |
| Subject Sample # 2 | 0.197 | 00:06 |
| Breath Volume      | 1.515 | 00:06 |
| Air Blank          | 0.000 | 00:06 |
| Control Test       | 0.079 | 00:07 |
| Air Blank          | 0.000 | 00:07 |
| Diagnostics Check  | OK    | 00:07 |

---

Data Download Date/Time: 04/28/2016 13:22

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FELIX,GEORGE,D  
BREATH TEST OPERATOR AGENCY: St. Lucie County SO  
ARRESTING OFFICER: FELIX,GEORGE  
ARRESTING OFFICER AGENCY: St. Lucie County SO  
LAST AGENCY INSPECTION: 03/16/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26614080A7  
DRY GAS EXPIRY DATE: 10/05/2016  
VIOLATION CODE: DUI  
ARREST TIME: 01:09  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:53

SUBJECT'S LAST NAME: GENT

DATE/TIME OF TEST: 04/17/2016 02:26:11

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 02:22 |                      |
| Air Blank         | 0.000 | 02:22 |                      |
| Control Test      | 0.080 | 02:23 |                      |
| Air Blank         | 0.000 | 02:23 |                      |
| Subject Sample    | # 1   | 02:24 | Subject Test Refused |
| Breath Volume     | 0.000 | 02:24 |                      |
| Air Blank         | 0.000 | 02:25 |                      |
| Control Test      | 0.080 | 02:25 |                      |
| Air Blank         | 0.000 | 02:26 |                      |
| Diagnostics Check | OK    | 02:26 |                      |

---

Data Download Date/Time: 04/28/2016 13:22

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ALVAREZ,JUAN,R  
BREATH TEST OPERATOR AGENCY: Fort Pierce PD  
ARRESTING OFFICER: WACHENDORFER,JEFF  
ARRESTING OFFICER AGENCY: Fort Pierce PD  
LAST AGENCY INSPECTION: 03/16/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26614080A7  
DRY GAS EXPIRY DATE: 10/05/2016  
VIOLATION CODE: DUI  
ARREST TIME: 23:21  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:01

SUBJECT'S LAST NAME: ONEAL

DATE/TIME OF TEST: 04/18/2016 23:46:06

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:39 |
| Air Blank          | 0.000 | 23:39 |
| Control Test       | 0.080 | 23:40 |
| Air Blank          | 0.000 | 23:40 |
| Subject Sample # 1 | 0.160 | 23:41 |
| Breath Volume      | 1.917 | 23:41 |
| Air Blank          | 0.000 | 23:42 |
| Air Blank          | 0.000 | 23:43 |
| Subject Sample # 2 | 0.156 | 23:44 |
| Breath Volume      | 1.984 | 23:44 |
| Air Blank          | 0.000 | 23:45 |
| Control Test       | 0.079 | 23:45 |
| Air Blank          | 0.000 | 23:46 |
| Diagnostics Check  | OK    | 23:46 |

---

Data Download Date/Time: 04/28/2016 13:22

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PAYNE,JAMES,B  
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)  
ARRESTING OFFICER: FRETWELL,JONATHAN  
ARRESTING OFFICER AGENCY: FFWCC (West Palm)  
LAST AGENCY INSPECTION: 03/16/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26614080A7  
DRY GAS EXPIRY DATE: 10/05/2016  
VIOLATION CODE: BUI  
ARREST TIME: 21:13  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:58

SUBJECT'S LAST NAME: VANDYGRIF

DATE/TIME OF TEST: 04/20/2016 22:24:58

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:18 |
| Air Blank          | 0.000 | 22:19 |
| Control Test       | 0.080 | 22:19 |
| Air Blank          | 0.000 | 22:19 |
| Subject Sample # 1 | 0.000 | 22:20 |
| Breath Volume      | 2.832 | 22:20 |
| Air Blank          | 0.000 | 22:21 |
| Air Blank          | 0.000 | 22:23 |
| Subject Sample # 2 | 0.007 | 22:23 |
| Breath Volume      | 3.148 | 22:23 |
| Air Blank          | 0.000 | 22:24 |
| Control Test       | 0.080 | 22:24 |
| Air Blank          | 0.000 | 22:24 |
| Diagnostics Check  | OK    | 22:24 |

---

Data Download Date/Time: 04/28/2016 13:22

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HOGAN,TODD,A  
BREATH TEST OPERATOR AGENCY: St. Lucie County SO  
ARRESTING OFFICER: HOGAN,TODD  
ARRESTING OFFICER AGENCY: St. Lucie County SO  
LAST AGENCY INSPECTION: 03/16/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26614080A7  
DRY GAS EXPIRY DATE: 10/05/2016  
VIOLATION CODE: DUI  
ARREST TIME: 02:24  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:04

SUBJECT'S LAST NAME: COFFMAN

DATE/TIME OF TEST: 04/21/2016 03:31:59

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:25 |
| Air Blank          | 0.000 | 03:26 |
| Control Test       | 0.080 | 03:26 |
| Air Blank          | 0.000 | 03:27 |
| Subject Sample # 1 | 0.269 | 03:27 |
| Breath Volume      | 2.500 | 03:27 |
| Air Blank          | 0.000 | 03:28 |
| Air Blank          | 0.000 | 03:30 |
| Subject Sample # 2 | 0.267 | 03:30 |
| Breath Volume      | 2.148 | 03:30 |
| Air Blank          | 0.000 | 03:31 |
| Control Test       | 0.079 | 03:31 |
| Air Blank          | 0.000 | 03:31 |
| Diagnostics Check  | OK    | 03:31 |

---

Data Download Date/Time: 04/28/2016 13:22

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HOGAN,TODD,A  
BREATH TEST OPERATOR AGENCY: St. Lucie County SO  
ARRESTING OFFICER: HOGAN,TODD  
ARRESTING OFFICER AGENCY: St. Lucie County SO  
LAST AGENCY INSPECTION: 03/16/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26614080A7  
DRY GAS EXPIRY DATE: 10/05/2016  
VIOLATION CODE: DUI  
ARREST TIME: 23:52  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: KING

DATE/TIME OF TEST: 04/23/2016 00:38:08

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:31 |
| Air Blank          | 0.000 | 00:32 |
| Control Test       | 0.080 | 00:32 |
| Air Blank          | 0.000 | 00:33 |
| Subject Sample # 1 | 0.152 | 00:33 |
| Breath Volume      | 2.160 | 00:33 |
| Air Blank          | 0.000 | 00:34 |
| Air Blank          | 0.000 | 00:36 |
| Subject Sample # 2 | 0.149 | 00:36 |
| Breath Volume      | 1.816 | 00:36 |
| Air Blank          | 0.000 | 00:37 |
| Control Test       | 0.080 | 00:37 |
| Air Blank          | 0.000 | 00:38 |
| Diagnostics Check  | OK    | 00:38 |

---

Data Download Date/Time: 04/28/2016 13:22

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BRENNAN,LAURIE,B  
BREATH TEST OPERATOR AGENCY: St. Lucie County SO  
ARRESTING OFFICER: BRENNAN,LAURIE  
ARRESTING OFFICER AGENCY: St. Lucie County SO  
LAST AGENCY INSPECTION: 03/16/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26614080A7  
DRY GAS EXPIRY DATE: 10/05/2016  
VIOLATION CODE: DUI  
ARREST TIME: 00:54  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:16

SUBJECT'S LAST NAME: ARIAS JIMENEZ

DATE/TIME OF TEST: 04/23/2016 01:54:50

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:46 |
| Air Blank          | 0.000 | 01:47 |
| Control Test       | 0.080 | 01:47 |
| Air Blank          | 0.000 | 01:48 |
| Subject Sample # 1 | 0.165 | 01:49 |
| Breath Volume      | 3.109 | 01:49 |
| Air Blank          | 0.000 | 01:50 |
| Air Blank          | 0.000 | 01:52 |
| Subject Sample # 2 | 0.152 | 01:53 |
| Breath Volume      | 3.156 | 01:53 |
| Air Blank          | 0.000 | 01:53 |
| Control Test       | 0.079 | 01:54 |
| Air Blank          | 0.000 | 01:54 |
| Diagnostics Check  | OK    | 01:54 |

---

Data Download Date/Time: 04/28/2016 13:22



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HOGAN,TODD,A  
BREATH TEST OPERATOR AGENCY: St. Lucie County SO  
ARRESTING OFFICER: HOGAN,TODD  
ARRESTING OFFICER AGENCY: St. Lucie County SO  
LAST AGENCY INSPECTION: 03/16/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26614080A7  
DRY GAS EXPIRY DATE: 10/05/2016  
VIOLATION CODE: DUI  
ARREST TIME: 23:49  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:13

SUBJECT'S LAST NAME: YARNELL

DATE/TIME OF TEST: 04/24/2016 00:42:33

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:36 |
| Air Blank          | 0.000 | 00:36 |
| Control Test       | 0.080 | 00:37 |
| Air Blank          | 0.000 | 00:37 |
| Subject Sample # 1 | 0.169 | 00:38 |
| Breath Volume      | 2.046 | 00:38 |
| Air Blank          | 0.000 | 00:38 |
| Air Blank          | 0.000 | 00:40 |
| Subject Sample # 2 | 0.170 | 00:41 |
| Breath Volume      | 1.796 | 00:41 |
| Air Blank          | 0.000 | 00:41 |
| Control Test       | 0.080 | 00:41 |
| Air Blank          | 0.000 | 00:42 |
| Diagnostics Check  | OK    | 00:42 |

---

Data Download Date/Time: 04/28/2016 13:22

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: STOKES,SHAYNE,R  
BREATH TEST OPERATOR AGENCY: Fort Pierce PD  
ARRESTING OFFICER: STOKES,SHAYNE  
ARRESTING OFFICER AGENCY: Fort Pierce PD  
LAST AGENCY INSPECTION: 03/16/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26614080A7  
DRY GAS EXPIRY DATE: 10/05/2016  
VIOLATION CODE: DUI  
ARREST TIME: 10:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 10:30

SUBJECT'S LAST NAME: PINEDA

DATE/TIME OF TEST: 04/24/2016 11:00:38

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 10:54 |
| Air Blank          | 0.000 | 10:54 |
| Control Test       | 0.079 | 10:55 |
| Air Blank          | 0.000 | 10:55 |
| Subject Sample # 1 | 0.000 | 10:56 |
| Breath Volume      | 1.917 | 10:56 |
| Air Blank          | 0.000 | 10:56 |
| Air Blank          | 0.000 | 10:58 |
| Subject Sample # 2 | 0.000 | 10:59 |
| Breath Volume      | 1.769 | 10:59 |
| Air Blank          | 0.000 | 10:59 |
| Control Test       | 0.079 | 11:00 |
| Air Blank          | 0.000 | 11:00 |
| Diagnostics Check  | OK    | 11:00 |

---

Data Download Date/Time: 04/28/2016 13:22

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ALVAREZ,JUAN,R  
BREATH TEST OPERATOR AGENCY: Fort Pierce PD  
ARRESTING OFFICER: MACCARTHY,KYLE  
ARRESTING OFFICER AGENCY: Fort Pierce PD  
LAST AGENCY INSPECTION: 03/16/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26614080A7  
DRY GAS EXPIRY DATE: 10/05/2016  
VIOLATION CODE: DUI  
ARREST TIME: 00:03  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:03

SUBJECT'S LAST NAME: RYAN

DATE/TIME OF TEST: 04/25/2016 00:44:32

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:37 |
| Air Blank          | 0.000 | 00:38 |
| Control Test       | 0.079 | 00:38 |
| Air Blank          | 0.000 | 00:38 |
| Subject Sample # 1 | 0.198 | 00:39 |
| Breath Volume      | 4.199 | 00:39 |
| Air Blank          | 0.000 | 00:40 |
| Air Blank          | 0.000 | 00:41 |
| Subject Sample # 2 | 0.205 | 00:43 |
| Breath Volume      | 4.507 | 00:43 |
| Air Blank          | 0.000 | 00:43 |
| Control Test       | 0.080 | 00:43 |
| Air Blank          | 0.000 | 00:44 |
| Diagnostics Check  | OK    | 00:44 |

---

Data Download Date/Time: 04/28/2016 13:22

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CAGLIONI, DAVID, A  
BREATH TEST OPERATOR AGENCY: St. Lucie County SO  
ARRESTING OFFICER: LOPEZ, EDWARD  
ARRESTING OFFICER AGENCY: St. Lucie County SO  
LAST AGENCY INSPECTION: 04/28/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24515080A1  
DRY GAS EXPIRY DATE: 10/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 15:13  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 16:15

SUBJECT'S LAST NAME: CELENTANO

DATE/TIME OF TEST: 04/29/2016 16:41:57

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 16:36 |
| Air Blank          | 0.000 | 16:36 |
| Control Test       | 0.080 | 16:36 |
| Air Blank          | 0.000 | 16:37 |
| Subject Sample # 1 | 0.183 | 16:37 |
| Breath Volume      | 1.609 | 16:37 |
| Air Blank          | 0.000 | 16:38 |
| Air Blank          | 0.000 | 16:40 |
| Subject Sample # 2 | 0.178 | 16:40 |
| Breath Volume      | 1.808 | 16:40 |
| Air Blank          | 0.000 | 16:41 |
| Control Test       | 0.080 | 16:41 |
| Air Blank          | 0.000 | 16:41 |
| Diagnostics Check  | OK    | 16:41 |

---

Data Download Date/Time: 05/21/2016 10:18

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PAYNE,JAMES,B  
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)  
ARRESTING OFFICER: ARNOLD,JON  
ARRESTING OFFICER AGENCY: FFWCC (West Palm)  
LAST AGENCY INSPECTION: 04/28/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24515080A1  
DRY GAS EXPIRY DATE: 10/05/2017  
VIOLATION CODE: BUI  
ARREST TIME: 12:33  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 12:54

SUBJECT'S LAST NAME: RAMOS

DATE/TIME OF TEST: 04/30/2016 13:27:06

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 13:20 |
| Air Blank          | 0.000 | 13:21 |
| Control Test       | 0.080 | 13:21 |
| Air Blank          | 0.000 | 13:22 |
| Subject Sample # 1 | 0.000 | 13:22 |
| Breath Volume      | 2.976 | 13:22 |
| Air Blank          | 0.000 | 13:23 |
| Air Blank          | 0.000 | 13:25 |
| Subject Sample # 2 | 0.000 | 13:25 |
| Breath Volume      | 2.914 | 13:25 |
| Air Blank          | 0.000 | 13:26 |
| Control Test       | 0.080 | 13:26 |
| Air Blank          | 0.000 | 13:27 |
| Diagnostics Check  | OK    | 13:27 |

---

Data Download Date/Time: 05/21/2016 10:18

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CAGLIONI,DAVID,A  
BREATH TEST OPERATOR AGENCY: St. Lucie County SO  
ARRESTING OFFICER: CAGLIONI,DAVID  
ARRESTING OFFICER AGENCY: St. Lucie County SO  
LAST AGENCY INSPECTION: 04/28/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24515080A1  
DRY GAS EXPIRY DATE: 10/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 15:42  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:10

SUBJECT'S LAST NAME: LYNCH

DATE/TIME OF TEST: 04/30/2016 16:36:25

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 16:30 |
| Air Blank          | 0.000 | 16:30 |
| Control Test       | 0.081 | 16:30 |
| Air Blank          | 0.000 | 16:31 |
| Subject Sample # 1 | 0.053 | 16:32 |
| Breath Volume      | 1.167 | 16:32 |
| Air Blank          | 0.000 | 16:32 |
| Air Blank          | 0.000 | 16:34 |
| Subject Sample # 2 | 0.058 | 16:34 |
| Breath Volume      | 1.210 | 16:34 |
| Air Blank          | 0.000 | 16:35 |
| Control Test       | 0.080 | 16:35 |
| Air Blank          | 0.000 | 16:36 |
| Diagnostics Check  | OK    | 16:36 |

---

Data Download Date/Time: 05/21/2016 10:18

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000795

INSTRUMENT REGISTERED TO: GULF COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HERNANDEZ,SONYA,C  
BREATH TEST OPERATOR AGENCY: Port St. Joe PD  
ARRESTING OFFICER: HERNANDEZ,SONYA  
ARRESTING OFFICER AGENCY: Port St. Joe PD  
LAST AGENCY INSPECTION: 03/31/2016

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 19615080A2  
DRY GAS EXPIRY DATE: 09/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 21:31  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:46

SUBJECT'S LAST NAME: BAXTER

DATE/TIME OF TEST: 04/01/2016 23:31:31

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:25 |
| Air Blank          | 0.000 | 23:25 |
| Control Test       | 0.076 | 23:26 |
| Air Blank          | 0.000 | 23:26 |
| Subject Sample # 1 | 0.041 | 23:26 |
| Breath Volume      | 3.707 | 23:26 |
| Air Blank          | 0.000 | 23:27 |
| Air Blank          | 0.000 | 23:29 |
| Subject Sample # 2 | 0.038 | 23:30 |
| Breath Volume      | 2.675 | 23:30 |
| Air Blank          | 0.000 | 23:30 |
| Control Test       | 0.077 | 23:30 |
| Air Blank          | 0.000 | 23:31 |
| Diagnostics Check  | OK    | 23:31 |

---

Data Download Date/Time: 04/28/2016 15:13

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TORRES,JENIFER,  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: DUNCAN,M  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03415080A2  
DRY GAS EXPIRY DATE: 03/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 21:47  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: RAMIREZ RAMIREZ

DATE/TIME OF TEST: 04/08/2016 22:47:42

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:41 |
| Air Blank          | 0.000 | 22:41 |
| Control Test       | 0.078 | 22:41 |
| Air Blank          | 0.000 | 22:42 |
| Subject Sample # 1 | 0.225 | 22:43 |
| Breath Volume      | 2.976 | 22:43 |
| Air Blank          | 0.000 | 22:43 |
| Air Blank          | 0.000 | 22:45 |
| Subject Sample # 2 | 0.228 | 22:46 |
| Breath Volume      | 2.875 | 22:46 |
| Air Blank          | 0.000 | 22:46 |
| Control Test       | 0.077 | 22:47 |
| Air Blank          | 0.000 | 22:47 |
| Diagnostics Check  | OK    | 22:47 |

---

Data Download Date/Time: 05/11/2016 10:16



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HENRY, PATRICIA,  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: VAN-TREESE, S  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03415080A2  
DRY GAS EXPIRY DATE: 03/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 02:08  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:32

SUBJECT'S LAST NAME: BRUNS

DATE/TIME OF TEST: 04/09/2016 03:03:24

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:52 |
| Air Blank          | 0.000 | 02:53 |
| Control Test       | 0.079 | 02:53 |
| Air Blank          | 0.000 | 02:53 |
| Subject Sample # 1 | 0.227 | 02:55 |
| Breath Volume      | 2.058 | 02:55 |
| Air Blank          | 0.000 | 02:55 |
| Air Blank          | 0.000 | 02:57 |
| Subject Sample # 2 | 0.261 | 02:58 |
| Breath Volume      | 1.988 | 02:58 |
| Air Blank          | 0.000 | 02:58 |
| Air Blank          | 0.000 | 03:00 |
| Subject Sample # 3 | 0.222 | 03:01 |
| Breath Volume      | 1.246 | 03:01 |
| Air Blank          | 0.000 | 03:02 |
| Control Test       | 0.077 | 03:02 |
| Air Blank          | 0.000 | 03:03 |
| Diagnostics Check  | OK    | 03:03 |

---

Data Download Date/Time: 05/11/2016 10:16

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SARRO,TAYLOR,C  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BAILEY,D  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03415080A2  
DRY GAS EXPIRY DATE: 03/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 04:23  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:47

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 04/09/2016 05:19:43

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 05:12 |
| Air Blank          | 0.000 | 05:13 |
| Control Test       | 0.077 | 05:13 |
| Air Blank          | 0.000 | 05:14 |
| Subject Sample # 1 | 0.211 | 05:15 |
| Breath Volume      | 1.933 | 05:15 |
| Air Blank          | 0.000 | 05:15 |
| Air Blank          | 0.000 | 05:17 |
| Subject Sample # 2 | 0.225 | 05:18 |
| Breath Volume      | 2.542 | 05:18 |
| Air Blank          | 0.000 | 05:18 |
| Control Test       | 0.078 | 05:19 |
| Air Blank          | 0.000 | 05:19 |
| Diagnostics Check  | OK    | 05:19 |

---

Data Download Date/Time: 05/11/2016 10:16

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MCMINN,KATHY,  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: MALONE,C  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03415080A2  
DRY GAS EXPIRY DATE: 03/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 20:11  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: SANDERS

DATE/TIME OF TEST: 04/09/2016 21:18:13

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 21:10 |
| Air Blank          | 0.000 | 21:11 |
| Control Test       | 0.078 | 21:11 |
| Air Blank          | 0.000 | 21:12 |
| Subject Sample # 1 | 0.108 | 21:13 |
| Breath Volume      | 1.539 | 21:13 |
| Air Blank          | 0.000 | 21:13 |
| Air Blank          | 0.000 | 21:15 |
| Subject Sample # 2 | 0.112 | 21:16 |
| Breath Volume      | 1.394 | 21:16 |
| Air Blank          | 0.000 | 21:17 |
| Control Test       | 0.078 | 21:17 |
| Air Blank          | 0.000 | 21:18 |
| Diagnostics Check  | OK    | 21:18 |

---

Data Download Date/Time: 05/11/2016 10:16

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MCMINN,KATHY,  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: WISE,V  
ARRESTING OFFICER AGENCY: Plant City PD  
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03415080A2  
DRY GAS EXPIRY DATE: 03/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 22:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:13

SUBJECT'S LAST NAME: JUAREZ-VELASCO

DATE/TIME OF TEST: 04/09/2016 23:41:35

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:34 |
| Air Blank          | 0.000 | 23:34 |
| Control Test       | 0.079 | 23:34 |
| Air Blank          | 0.000 | 23:35 |
| Subject Sample # 1 | 0.242 | 23:36 |
| Breath Volume      | 2.656 | 23:36 |
| Air Blank          | 0.000 | 23:36 |
| Air Blank          | 0.000 | 23:38 |
| Subject Sample # 2 | 0.234 | 23:40 |
| Breath Volume      | 2.453 | 23:40 |
| Air Blank          | 0.000 | 23:40 |
| Control Test       | 0.078 | 23:40 |
| Air Blank          | 0.000 | 23:41 |
| Diagnostics Check  | OK    | 23:41 |

---

Data Download Date/Time: 05/11/2016 10:16

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KIRK,CASEY,S  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BAILEY,D  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03415080A2  
DRY GAS EXPIRY DATE: 03/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 02:36  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: SALINAS-HERNANDEZ

DATE/TIME OF TEST: 04/10/2016 03:36:54

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:30 |
| Air Blank          | 0.000 | 03:31 |
| Control Test       | 0.079 | 03:31 |
| Air Blank          | 0.000 | 03:31 |
| Subject Sample # 1 | 0.178 | 03:32 |
| Breath Volume      | 3.621 | 03:32 |
| Air Blank          | 0.000 | 03:33 |
| Air Blank          | 0.000 | 03:34 |
| Subject Sample # 2 | 0.176 | 03:35 |
| Breath Volume      | 2.937 | 03:35 |
| Air Blank          | 0.000 | 03:35 |
| Control Test       | 0.078 | 03:36 |
| Air Blank          | 0.000 | 03:36 |
| Diagnostics Check  | OK    | 03:36 |

---

Data Download Date/Time: 05/11/2016 10:16

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KIRK,CASEY,S  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: SIMPKINS,D  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03415080A2  
DRY GAS EXPIRY DATE: 03/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 21:30  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:53

SUBJECT'S LAST NAME: MCDANIEL

DATE/TIME OF TEST: 04/10/2016 22:21:12

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:14 |
| Air Blank          | 0.000 | 22:15 |
| Control Test       | 0.079 | 22:15 |
| Air Blank          | 0.000 | 22:16 |
| Subject Sample # 1 | 0.144 | 22:16 |
| Breath Volume      | 1.824 | 22:16 |
| Air Blank          | 0.000 | 22:17 |
| Air Blank          | 0.000 | 22:19 |
| Subject Sample # 2 | 0.148 | 22:19 |
| Breath Volume      | 1.503 | 22:19 |
| Air Blank          | 0.000 | 22:20 |
| Control Test       | 0.078 | 22:20 |
| Air Blank          | 0.000 | 22:21 |
| Diagnostics Check  | OK    | 22:21 |

---

Data Download Date/Time: 05/11/2016 10:16

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MCMINN,KATHY,  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: VAN-TREESE,S  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03415080A2  
DRY GAS EXPIRY DATE: 03/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 00:56  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:22

SUBJECT'S LAST NAME: LAMAR

DATE/TIME OF TEST: 04/12/2016 01:50:01

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:42 |
| Air Blank          | 0.000 | 01:43 |
| Control Test       | 0.079 | 01:43 |
| Air Blank          | 0.000 | 01:44 |
| Subject Sample # 1 | 0.176 | 01:45 |
| Breath Volume      | 1.683 | 01:45 |
| Air Blank          | 0.000 | 01:45 |
| Air Blank          | 0.000 | 01:47 |
| Subject Sample # 2 | 0.174 | 01:48 |
| Breath Volume      | 1.593 | 01:48 |
| Air Blank          | 0.000 | 01:49 |
| Control Test       | 0.078 | 01:49 |
| Air Blank          | 0.000 | 01:49 |
| Diagnostics Check  | OK    | 01:50 |

---

Data Download Date/Time: 05/11/2016 10:16

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BRIGHTWELL,R  
ARRESTING OFFICER AGENCY: U of So FL Campus PD  
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03415080A2  
DRY GAS EXPIRY DATE: 03/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 04:09  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:48

SUBJECT'S LAST NAME: FEBRES

DATE/TIME OF TEST: 04/12/2016 05:16:45

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 05:09 |
| Air Blank          | 0.000 | 05:09 |
| Control Test       | 0.078 | 05:09 |
| Air Blank          | 0.000 | 05:10 |
| Subject Sample # 1 | 0.152 | 05:12 |
| Breath Volume      | 2.957 | 05:12 |
| Air Blank          | 0.000 | 05:12 |
| Air Blank          | 0.000 | 05:14 |
| Subject Sample # 2 | 0.156 | 05:15 |
| Breath Volume      | 1.679 | 05:15 |
| Air Blank          | 0.000 | 05:15 |
| Control Test       | 0.077 | 05:16 |
| Air Blank          | 0.000 | 05:16 |
| Diagnostics Check  | OK    | 05:16 |

---

Data Download Date/Time: 05/11/2016 10:16



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SARRO,TAYLOR,C  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: AGUAYO,J  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03415080A2  
DRY GAS EXPIRY DATE: 03/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 02:06  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:28

SUBJECT'S LAST NAME: POPE

DATE/TIME OF TEST: 04/14/2016 02:55:24

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:48 |
| Air Blank          | 0.000 | 02:49 |
| Control Test       | 0.078 | 02:49 |
| Air Blank          | 0.000 | 02:50 |
| Subject Sample # 1 | 0.188 | 02:50 |
| Breath Volume      | 2.449 | 02:50 |
| Air Blank          | 0.000 | 02:51 |
| Air Blank          | 0.000 | 02:53 |
| Subject Sample # 2 | 0.195 | 02:53 |
| Breath Volume      | 1.859 | 02:53 |
| Air Blank          | 0.000 | 02:54 |
| Control Test       | 0.078 | 02:54 |
| Air Blank          | 0.000 | 02:55 |
| Diagnostics Check  | OK    | 02:55 |

---

Data Download Date/Time: 05/11/2016 10:16

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TORRES,JENIFER,  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: VISSER,A  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03415080A2  
DRY GAS EXPIRY DATE: 03/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 03:04  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:27

SUBJECT'S LAST NAME: BRUZUAL SEPULVEDA

DATE/TIME OF TEST: 04/14/2016 03:51:29

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 03:48 |                      |
| Air Blank         | 0.000 | 03:48 |                      |
| Control Test      | 0.078 | 03:48 |                      |
| Air Blank         | 0.000 | 03:49 |                      |
| Subject Sample    | # 1   | 03:50 | Subject Test Refused |
| Breath Volume     | 0.000 | 03:50 |                      |
| Air Blank         | 0.000 | 03:50 |                      |
| Control Test      | 0.078 | 03:50 |                      |
| Air Blank         | 0.000 | 03:51 |                      |
| Diagnostics Check | OK    | 03:51 |                      |

---

Data Download Date/Time: 05/11/2016 10:16

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: SIMPKINS,D  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03415080A2  
DRY GAS EXPIRY DATE: 03/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 22:01  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:01

SUBJECT'S LAST NAME: ROSENBLUM

DATE/TIME OF TEST: 04/14/2016 23:28:16

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:21 |
| Air Blank          | 0.000 | 23:22 |
| Control Test       | 0.078 | 23:22 |
| Air Blank          | 0.000 | 23:23 |
| Subject Sample # 1 | 0.000 | 23:23 |
| Breath Volume      | 1.335 | 23:23 |
| Air Blank          | 0.000 | 23:24 |
| Air Blank          | 0.000 | 23:26 |
| Subject Sample # 2 | 0.000 | 23:26 |
| Breath Volume      | 1.828 | 23:26 |
| Air Blank          | 0.000 | 23:27 |
| Control Test       | 0.078 | 23:27 |
| Air Blank          | 0.000 | 23:28 |
| Diagnostics Check  | OK    | 23:28 |

---

Data Download Date/Time: 05/11/2016 10:16

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MCMINN,KATHY,  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: GONZALEZ,J  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03415080A2  
DRY GAS EXPIRY DATE: 03/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 00:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:04

SUBJECT'S LAST NAME: KOENIG

DATE/TIME OF TEST: 04/15/2016 01:31:29

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:24 |
| Air Blank          | 0.000 | 01:25 |
| Control Test       | 0.078 | 01:25 |
| Air Blank          | 0.000 | 01:26 |
| Subject Sample # 1 | 0.028 | 01:26 |
| Breath Volume      | 3.750 | 01:26 |
| Air Blank          | 0.000 | 01:27 |
| Air Blank          | 0.000 | 01:29 |
| Subject Sample # 2 | 0.027 | 01:29 |
| Breath Volume      | 2.988 | 01:29 |
| Air Blank          | 0.000 | 01:30 |
| Control Test       | 0.079 | 01:30 |
| Air Blank          | 0.000 | 01:31 |
| Diagnostics Check  | OK    | 01:31 |

---

Data Download Date/Time: 05/11/2016 10:16

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: MALONE,C  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03415080A2  
DRY GAS EXPIRY DATE: 03/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 03:07  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:58

SUBJECT'S LAST NAME: GILBERT

DATE/TIME OF TEST: 04/15/2016 04:25:24

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:18 |
| Air Blank          | 0.000 | 04:19 |
| Control Test       | 0.079 | 04:19 |
| Air Blank          | 0.000 | 04:20 |
| Subject Sample # 1 | 0.000 | 04:21 |
| Breath Volume      | 1.367 | 04:21 |
| Air Blank          | 0.000 | 04:21 |
| Air Blank          | 0.000 | 04:23 |
| Subject Sample # 2 | 0.000 | 04:24 |
| Breath Volume      | 2.347 | 04:24 |
| Air Blank          | 0.000 | 04:24 |
| Control Test       | 0.078 | 04:24 |
| Air Blank          | 0.000 | 04:25 |
| Diagnostics Check  | OK    | 04:25 |

---

Data Download Date/Time: 05/11/2016 10:16

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: SNAGGS,S  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03415080A2  
DRY GAS EXPIRY DATE: 03/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 22:08  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:44

SUBJECT'S LAST NAME: SIMPSON

DATE/TIME OF TEST: 04/15/2016 23:11:53

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:05 |
| Air Blank          | 0.000 | 23:05 |
| Control Test       | 0.079 | 23:05 |
| Air Blank          | 0.000 | 23:06 |
| Subject Sample # 1 | 0.180 | 23:07 |
| Breath Volume      | 2.828 | 23:07 |
| Air Blank          | 0.000 | 23:07 |
| Air Blank          | 0.000 | 23:09 |
| Subject Sample # 2 | 0.178 | 23:10 |
| Breath Volume      | 2.988 | 23:10 |
| Air Blank          | 0.000 | 23:10 |
| Control Test       | 0.078 | 23:11 |
| Air Blank          | 0.000 | 23:11 |
| Diagnostics Check  | OK    | 23:11 |

---

Data Download Date/Time: 05/11/2016 10:16

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MCMINN,KATHY,  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: PALFY,R  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03415080A2  
DRY GAS EXPIRY DATE: 03/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 00:33  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:21

SUBJECT'S LAST NAME: ORTIZ JR

DATE/TIME OF TEST: 04/16/2016 01:50:30

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:42 |
| Air Blank          | 0.000 | 01:42 |
| Control Test       | 0.078 | 01:43 |
| Air Blank          | 0.000 | 01:43 |
| Subject Sample # 1 | 0.107 | 01:45 |
| Breath Volume      | 1.738 | 01:45 |
| Air Blank          | 0.000 | 01:46 |
| Air Blank          | 0.000 | 01:48 |
| Subject Sample # 2 | 0.107 | 01:48 |
| Breath Volume      | 1.945 | 01:48 |
| Air Blank          | 0.000 | 01:49 |
| Control Test       | 0.077 | 01:49 |
| Air Blank          | 0.000 | 01:50 |
| Diagnostics Check  | OK    | 01:50 |

---

Data Download Date/Time: 05/11/2016 10:16

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: VISSER,A  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03415080A2  
DRY GAS EXPIRY DATE: 03/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 03:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: BRISSOLESE

DATE/TIME OF TEST: 04/16/2016 04:32:11

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:25 |
| Air Blank          | 0.000 | 04:25 |
| Control Test       | 0.078 | 04:26 |
| Air Blank          | 0.000 | 04:26 |
| Subject Sample # 1 | 0.143 | 04:27 |
| Breath Volume      | 1.816 | 04:27 |
| Air Blank          | 0.000 | 04:28 |
| Air Blank          | 0.000 | 04:29 |
| Subject Sample # 2 | 0.142 | 04:30 |
| Breath Volume      | 1.187 | 04:30 |
| Air Blank          | 0.000 | 04:31 |
| Control Test       | 0.077 | 04:31 |
| Air Blank          | 0.000 | 04:32 |
| Diagnostics Check  | OK    | 04:32 |

---

Data Download Date/Time: 05/11/2016 10:16



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MCMINN,KATHY,  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BAILEY,D  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03415080A2  
DRY GAS EXPIRY DATE: 03/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 04:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:26

SUBJECT'S LAST NAME: NORIEGA

DATE/TIME OF TEST: 04/16/2016 04:53:37

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:46 |
| Air Blank          | 0.000 | 04:46 |
| Control Test       | 0.078 | 04:47 |
| Air Blank          | 0.000 | 04:47 |
| Subject Sample # 1 | 0.152 | 04:48 |
| Breath Volume      | 1.125 | 04:48 |
| Air Blank          | 0.000 | 04:49 |
| Air Blank          | 0.000 | 04:50 |
| Subject Sample # 2 | 0.158 | 04:51 |
| Breath Volume      | 1.886 | 04:51 |
| Air Blank          | 0.000 | 04:52 |
| Control Test       | 0.077 | 04:52 |
| Air Blank          | 0.000 | 04:53 |
| Diagnostics Check  | OK    | 04:53 |

---

Data Download Date/Time: 05/11/2016 10:16

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SIMMONS,ANGELA,V  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: CABALLERO,C  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03415080A2  
DRY GAS EXPIRY DATE: 03/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 11:43  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 13:23

SUBJECT'S LAST NAME: DOLAN

DATE/TIME OF TEST: 04/16/2016 13:51:12

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 13:43 |
| Air Blank          | 0.000 | 13:44 |
| Control Test       | 0.078 | 13:44 |
| Air Blank          | 0.000 | 13:44 |
| Subject Sample # 1 | 0.234 | 13:45 |
| Breath Volume      | 2.113 | 13:45 |
| Air Blank          | 0.000 | 13:46 |
| Air Blank          | 0.000 | 13:47 |
| Subject Sample # 2 | 0.214 | 13:49 |
| Breath Volume      | 1.562 | 13:49 |
| Air Blank          | 0.000 | 13:50 |
| Control Test       | 0.078 | 13:50 |
| Air Blank          | 0.000 | 13:51 |
| Diagnostics Check  | OK    | 13:51 |

---

Data Download Date/Time: 05/11/2016 10:16

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BARLAUG, G  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03415080A2  
DRY GAS EXPIRY DATE: 03/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 19:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:44

SUBJECT'S LAST NAME: ALLEN

DATE/TIME OF TEST: 04/16/2016 20:10:59

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 20:04 |
| Air Blank          | 0.000 | 20:05 |
| Control Test       | 0.078 | 20:05 |
| Air Blank          | 0.000 | 20:06 |
| Subject Sample # 1 | 0.152 | 20:06 |
| Breath Volume      | 2.257 | 20:06 |
| Air Blank          | 0.000 | 20:07 |
| Air Blank          | 0.000 | 20:09 |
| Subject Sample # 2 | 0.154 | 20:09 |
| Breath Volume      | 2.878 | 20:09 |
| Air Blank          | 0.000 | 20:10 |
| Control Test       | 0.078 | 20:10 |
| Air Blank          | 0.000 | 20:10 |
| Diagnostics Check  | OK    | 20:10 |

---

Data Download Date/Time: 05/11/2016 10:16

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TORRES,JENIFER,  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BRABAND,M  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03415080A2  
DRY GAS EXPIRY DATE: 03/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 01:28  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:54

SUBJECT'S LAST NAME: KOENIG

DATE/TIME OF TEST: 04/17/2016 02:20:49

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:14 |
| Air Blank          | 0.000 | 02:15 |
| Control Test       | 0.079 | 02:15 |
| Air Blank          | 0.000 | 02:15 |
| Subject Sample # 1 | 0.112 | 02:16 |
| Breath Volume      | 1.128 | 02:16 |
| Air Blank          | 0.000 | 02:16 |
| Air Blank          | 0.000 | 02:18 |
| Subject Sample # 2 | 0.122 | 02:19 |
| Breath Volume      | 1.164 | 02:19 |
| Air Blank          | 0.000 | 02:19 |
| Control Test       | 0.077 | 02:20 |
| Air Blank          | 0.000 | 02:20 |
| Diagnostics Check  | OK    | 02:20 |

---

Data Download Date/Time: 05/11/2016 10:16

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HENRY, PATRICIA,  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: VAN-TREESE, S  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03415080A2  
DRY GAS EXPIRY DATE: 03/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 02:39  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 04/17/2016 03:32:38

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:25 |
| Air Blank          | 0.000 | 03:25 |
| Control Test       | 0.078 | 03:26 |
| Air Blank          | 0.000 | 03:26 |
| Subject Sample # 1 | 0.119 | 03:28 |
| Breath Volume      | 1.980 | 03:28 |
| Air Blank          | 0.000 | 03:28 |
| Air Blank          | 0.000 | 03:30 |
| Subject Sample # 2 | 0.120 | 03:31 |
| Breath Volume      | 1.441 | 03:31 |
| Air Blank          | 0.000 | 03:31 |
| Control Test       | 0.077 | 03:32 |
| Air Blank          | 0.000 | 03:32 |
| Diagnostics Check  | OK    | 03:32 |

---

Data Download Date/Time: 05/11/2016 10:16

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SARRO,TAYLOR,C  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: MINAYA,J  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03415080A2  
DRY GAS EXPIRY DATE: 03/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 03:18  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:02

SUBJECT'S LAST NAME: ALVAREZ AGUILAR

DATE/TIME OF TEST: 04/17/2016 04:28:47

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:22 |
| Air Blank          | 0.000 | 04:22 |
| Control Test       | 0.079 | 04:23 |
| Air Blank          | 0.000 | 04:23 |
| Subject Sample # 1 | 0.111 | 04:24 |
| Breath Volume      | 1.101 | 04:24 |
| Air Blank          | 0.000 | 04:25 |
| Air Blank          | 0.000 | 04:26 |
| Subject Sample # 2 | 0.115 | 04:27 |
| Breath Volume      | 1.671 | 04:27 |
| Air Blank          | 0.000 | 04:27 |
| Control Test       | 0.077 | 04:28 |
| Air Blank          | 0.000 | 04:28 |
| Diagnostics Check  | OK    | 04:28 |

---

Data Download Date/Time: 05/11/2016 10:16

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BARLAUG, G  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03415080A2  
DRY GAS EXPIRY DATE: 03/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 04:31  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:56

SUBJECT'S LAST NAME: JOLLY IV

DATE/TIME OF TEST: 04/17/2016 05:26:39

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 05:17 |
| Air Blank          | 0.000 | 05:17 |
| Control Test       | 0.079 | 05:17 |
| Air Blank          | 0.000 | 05:18 |
| Subject Sample # 1 | 0.152 | 05:20 |
| Breath Volume      | 2.957 | 05:20 |
| Air Blank          | 0.000 | 05:21 |
| Air Blank          | 0.000 | 05:23 |
| Subject Sample # 2 | 0.144 | 05:25 |
| Breath Volume      | 1.785 | 05:25 |
| Air Blank          | 0.000 | 05:25 |
| Control Test       | 0.077 | 05:26 |
| Air Blank          | 0.000 | 05:26 |
| Diagnostics Check  | OK    | 05:26 |

---

Data Download Date/Time: 05/11/2016 10:16

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HENRY,PATRICIA,  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: VISSER,A  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03415080A2  
DRY GAS EXPIRY DATE: 03/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 05:19  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:47

SUBJECT'S LAST NAME: AVILA-JIMENEZ

DATE/TIME OF TEST: 04/17/2016 06:14:04

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 06:07 |
| Air Blank          | 0.000 | 06:08 |
| Control Test       | 0.078 | 06:08 |
| Air Blank          | 0.000 | 06:08 |
| Subject Sample # 1 | 0.192 | 06:09 |
| Breath Volume      | 1.156 | 06:09 |
| Air Blank          | 0.000 | 06:10 |
| Air Blank          | 0.000 | 06:11 |
| Subject Sample # 2 | 0.198 | 06:12 |
| Breath Volume      | 1.804 | 06:12 |
| Air Blank          | 0.000 | 06:13 |
| Control Test       | 0.077 | 06:13 |
| Air Blank          | 0.000 | 06:13 |
| Diagnostics Check  | OK    | 06:14 |

---

Data Download Date/Time: 05/11/2016 10:17



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SARRO,TAYLOR,C  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: VAN-TREESE,S  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03415080A2  
DRY GAS EXPIRY DATE: 03/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 20:42  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:06

SUBJECT'S LAST NAME: VEGA

DATE/TIME OF TEST: 04/17/2016 21:35:55

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 21:26 |
| Air Blank          | 0.000 | 21:26 |
| Control Test       | 0.079 | 21:27 |
| Air Blank          | 0.000 | 21:27 |
| Subject Sample # 1 | 0.101 | 21:30 |
| Breath Volume      | 1.550 | 21:30 |
| Air Blank          | 0.000 | 21:30 |
| Air Blank          | 0.000 | 21:32 |
| Subject Sample # 2 | 0.105 | 21:34 |
| Breath Volume      | 1.601 | 21:34 |
| Air Blank          | 0.000 | 21:34 |
| Control Test       | 0.078 | 21:35 |
| Air Blank          | 0.000 | 21:35 |
| Diagnostics Check  | OK    | 21:35 |

---

Data Download Date/Time: 05/11/2016 10:17

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TORRES,JENIFER,  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: FRANKLIN,R  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03415080A2  
DRY GAS EXPIRY DATE: 03/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 02:38  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: REYES-CRUZ

DATE/TIME OF TEST: 04/18/2016 03:27:33

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:20 |
| Air Blank          | 0.000 | 03:21 |
| Control Test       | 0.079 | 03:21 |
| Air Blank          | 0.000 | 03:22 |
| Subject Sample # 1 | 0.127 | 03:22 |
| Breath Volume      | 3.433 | 03:22 |
| Air Blank          | 0.000 | 03:23 |
| Air Blank          | 0.000 | 03:25 |
| Subject Sample # 2 | 0.125 | 03:25 |
| Breath Volume      | 2.273 | 03:25 |
| Air Blank          | 0.000 | 03:26 |
| Control Test       | 0.078 | 03:26 |
| Air Blank          | 0.000 | 03:27 |
| Diagnostics Check  | OK    | 03:27 |

---

Data Download Date/Time: 05/11/2016 10:17

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HENRY,PATRICIA,  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BAILEY,D  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03415080A2  
DRY GAS EXPIRY DATE: 03/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 02:56  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:19

SUBJECT'S LAST NAME: LUCAS

DATE/TIME OF TEST: 04/18/2016 03:50:39

---

|                   |       |       |                    |
|-------------------|-------|-------|--------------------|
| Diagnostics Check | OK    | 03:39 |                    |
| Air Blank         | 0.000 | 03:39 |                    |
| Control Test      | 0.079 | 03:40 |                    |
| Air Blank         | 0.000 | 03:40 |                    |
| Subject Sample    | # 1   | 03:43 | No Sample Provided |
| Breath Volume     | 0.000 | 03:43 |                    |
| Air Blank         | 0.000 | 03:44 |                    |
| Air Blank         | 0.000 | 03:46 |                    |
| Subject Sample    | # 2   | 03:49 | No Sample Provided |
| Breath Volume     | 0.000 | 03:49 |                    |
| Air Blank         | 0.000 | 03:49 |                    |
| Control Test      | 0.079 | 03:50 |                    |
| Air Blank         | 0.000 | 03:50 |                    |
| Diagnostics Check | OK    | 03:50 |                    |

---

Data Download Date/Time: 05/11/2016 10:17

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HENRY,PATRICIA,  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BAILEY,D  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03415080A2  
DRY GAS EXPIRY DATE: 03/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 02:56  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:19

SUBJECT'S LAST NAME: LUCAS

DATE/TIME OF TEST: 04/18/2016 04:12:19

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:58 |
| Air Blank          | 0.000 | 03:59 |
| Control Test       | 0.078 | 03:59 |
| Air Blank          | 0.000 | 04:00 |
| Subject Sample # 1 | 0.178 | 04:02 |
| Breath Volume      | 1.847 | 04:02 |
| Air Blank          | 0.000 | 04:02 |
| Air Blank          | 0.000 | 04:04 |
| Subject Sample # 2 | 0.203 | 04:05 |
| Breath Volume      | 2.976 | 04:05 |
| Air Blank          | 0.000 | 04:06 |
| Air Blank          | 0.000 | 04:08 |
| Subject Sample # 3 | 0.188 | 04:10 |
| Breath Volume      | 1.742 | 04:10 |
| Air Blank          | 0.000 | 04:11 |
| Control Test       | 0.078 | 04:11 |
| Air Blank          | 0.000 | 04:12 |
| Diagnostics Check  | OK    | 04:12 |

---

Data Download Date/Time: 05/11/2016 10:17

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HENRY, PATRICIA,  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: UNO, D  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03415080A2  
DRY GAS EXPIRY DATE: 03/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 18:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:07

SUBJECT'S LAST NAME: MONACO

DATE/TIME OF TEST: 04/18/2016 19:44:00

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 19:38 |
| Air Blank          | 0.000 | 19:38 |
| Control Test       | 0.079 | 19:38 |
| Air Blank          | 0.000 | 19:39 |
| Subject Sample # 1 | 0.000 | 19:39 |
| Breath Volume      | 1.402 | 19:39 |
| Air Blank          | 0.000 | 19:40 |
| Air Blank          | 0.000 | 19:42 |
| Subject Sample # 2 | 0.000 | 19:42 |
| Breath Volume      | 1.242 | 19:42 |
| Air Blank          | 0.000 | 19:43 |
| Control Test       | 0.079 | 19:43 |
| Air Blank          | 0.000 | 19:43 |
| Diagnostics Check  | OK    | 19:44 |

---

Data Download Date/Time: 05/11/2016 10:19

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SARRO,TAYLOR,C  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: VAN-TREESE,S  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03415080A2  
DRY GAS EXPIRY DATE: 03/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 22:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:38

SUBJECT'S LAST NAME: ZINNER

DATE/TIME OF TEST: 04/18/2016 23:03:42

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 22:58 |                      |
| Air Blank         | 0.000 | 22:59 |                      |
| Control Test      | 0.079 | 22:59 |                      |
| Air Blank         | 0.000 | 22:59 |                      |
| Subject Sample    | # 1   | 23:02 | Subject Test Refused |
| Breath Volume     | 0.000 | 23:02 |                      |
| Air Blank         | 0.000 | 23:02 |                      |
| Control Test      | 0.078 | 23:03 |                      |
| Air Blank         | 0.000 | 23:03 |                      |
| Diagnostics Check | OK    | 23:03 |                      |

---

Data Download Date/Time: 05/11/2016 10:19

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TORRES,JENIFER,  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BAILEY,D  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03415080A2  
DRY GAS EXPIRY DATE: 03/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 04:00  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: STEVENS

DATE/TIME OF TEST: 04/19/2016 04:46:45

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:40 |
| Air Blank          | 0.000 | 04:40 |
| Control Test       | 0.079 | 04:41 |
| Air Blank          | 0.000 | 04:41 |
| Subject Sample # 1 | 0.095 | 04:42 |
| Breath Volume      | 2.515 | 04:42 |
| Air Blank          | 0.000 | 04:42 |
| Air Blank          | 0.000 | 04:44 |
| Subject Sample # 2 | 0.095 | 04:45 |
| Breath Volume      | 2.492 | 04:45 |
| Air Blank          | 0.000 | 04:45 |
| Control Test       | 0.078 | 04:46 |
| Air Blank          | 0.000 | 04:46 |
| Diagnostics Check  | OK    | 04:46 |

---

Data Download Date/Time: 05/11/2016 10:19

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: MALONE, C  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03415080A2  
DRY GAS EXPIRY DATE: 03/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 04:34  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:04

SUBJECT'S LAST NAME: CHILAK

DATE/TIME OF TEST: 04/19/2016 05:33:03

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 05:27 |
| Air Blank          | 0.000 | 05:27 |
| Control Test       | 0.079 | 05:28 |
| Air Blank          | 0.000 | 05:28 |
| Subject Sample # 1 | 0.000 | 05:28 |
| Breath Volume      | 2.386 | 05:28 |
| Air Blank          | 0.000 | 05:29 |
| Air Blank          | 0.000 | 05:31 |
| Subject Sample # 2 | 0.000 | 05:31 |
| Breath Volume      | 2.812 | 05:31 |
| Air Blank          | 0.000 | 05:32 |
| Control Test       | 0.078 | 05:32 |
| Air Blank          | 0.000 | 05:32 |
| Diagnostics Check  | OK    | 05:33 |

---

Data Download Date/Time: 05/11/2016 10:19



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KIRK,CASEY,S  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BARLAUG,G  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03415080A2  
DRY GAS EXPIRY DATE: 03/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 03:34  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:57

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 04/20/2016 04:32:04

---

|                   |           |       |                    |
|-------------------|-----------|-------|--------------------|
| Diagnostics Check | OK        | 04:17 |                    |
| Air Blank         | 0.000     | 04:17 |                    |
| Control Test      | 0.079     | 04:18 |                    |
| Air Blank         | 0.000     | 04:18 |                    |
| Subject Sample    | # 1 0.145 | 04:19 |                    |
| Breath Volume     | 1.984     | 04:19 |                    |
| Air Blank         | 0.000     | 04:20 |                    |
| Air Blank         | 0.000     | 04:21 |                    |
| Subject Sample    | # 2 0.120 | 04:25 | Volume Not Met     |
| Breath Volume     | 0.781     | 04:25 |                    |
| Air Blank         | 0.000     | 04:25 |                    |
| Air Blank         | 0.000     | 04:27 |                    |
| Subject Sample    | # 3       | 04:30 | No Sample Provided |
| Breath Volume     | 0.000     | 04:30 |                    |
| Air Blank         | 0.000     | 04:31 |                    |
| Control Test      | 0.077     | 04:31 |                    |
| Air Blank         | 0.000     | 04:31 |                    |
| Diagnostics Check | OK        | 04:32 |                    |

---

Data Download Date/Time: 05/11/2016 10:19

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MCMINN,KATHY,  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: MATAS,T  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03415080A2  
DRY GAS EXPIRY DATE: 03/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 17:53  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 18:39

SUBJECT'S LAST NAME: WALDECK

DATE/TIME OF TEST: 04/20/2016 19:07:25

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 18:59 |
| Air Blank          | 0.000 | 18:59 |
| Control Test       | 0.079 | 19:00 |
| Air Blank          | 0.000 | 19:00 |
| Subject Sample # 1 | 0.164 | 19:02 |
| Breath Volume      | 1.191 | 19:02 |
| Air Blank          | 0.000 | 19:03 |
| Air Blank          | 0.000 | 19:04 |
| Subject Sample # 2 | 0.161 | 19:05 |
| Breath Volume      | 1.574 | 19:05 |
| Air Blank          | 0.000 | 19:06 |
| Control Test       | 0.078 | 19:06 |
| Air Blank          | 0.000 | 19:07 |
| Diagnostics Check  | OK    | 19:07 |

---

Data Download Date/Time: 05/11/2016 10:19

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KIRK,CASEY,S  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: GONZALEZ,J  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03415080A2  
DRY GAS EXPIRY DATE: 03/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 00:13  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:58

SUBJECT'S LAST NAME: COLLAZO

DATE/TIME OF TEST: 04/21/2016 01:28:57

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:19 |
| Air Blank          | 0.000 | 01:20 |
| Control Test       | 0.078 | 01:20 |
| Air Blank          | 0.000 | 01:21 |
| Subject Sample # 1 | 0.151 | 01:23 |
| Breath Volume      | 1.296 | 01:23 |
| Air Blank          | 0.000 | 01:23 |
| Air Blank          | 0.000 | 01:25 |
| Subject Sample # 2 | 0.141 | 01:27 |
| Breath Volume      | 2.140 | 01:27 |
| Air Blank          | 0.000 | 01:28 |
| Control Test       | 0.078 | 01:28 |
| Air Blank          | 0.000 | 01:28 |
| Diagnostics Check  | OK    | 01:28 |

---

Data Download Date/Time: 05/11/2016 10:19

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MCMINN,KATHY,  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: VISSER,A  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03415080A2  
DRY GAS EXPIRY DATE: 03/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 02:29  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: PAYNE

DATE/TIME OF TEST: 04/21/2016 03:26:26

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:19 |
| Air Blank          | 0.000 | 03:20 |
| Control Test       | 0.078 | 03:20 |
| Air Blank          | 0.000 | 03:21 |
| Subject Sample # 1 | 0.000 | 03:21 |
| Breath Volume      | 1.566 | 03:21 |
| Air Blank          | 0.000 | 03:22 |
| Air Blank          | 0.000 | 03:24 |
| Subject Sample # 2 | 0.000 | 03:25 |
| Breath Volume      | 1.796 | 03:25 |
| Air Blank          | 0.000 | 03:25 |
| Control Test       | 0.078 | 03:25 |
| Air Blank          | 0.000 | 03:26 |
| Diagnostics Check  | OK    | 03:26 |

---

Data Download Date/Time: 05/11/2016 10:19

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HENRY,PATRICIA,  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: DAWSY,B  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03415080A2  
DRY GAS EXPIRY DATE: 03/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 18:11  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 18:36

SUBJECT'S LAST NAME: ALBA

DATE/TIME OF TEST: 04/21/2016 19:03:15

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 18:56 |
| Air Blank          | 0.000 | 18:57 |
| Control Test       | 0.079 | 18:57 |
| Air Blank          | 0.000 | 18:58 |
| Subject Sample # 1 | 0.135 | 18:58 |
| Breath Volume      | 1.378 | 18:58 |
| Air Blank          | 0.000 | 18:59 |
| Air Blank          | 0.000 | 19:01 |
| Subject Sample # 2 | 0.138 | 19:01 |
| Breath Volume      | 1.585 | 19:01 |
| Air Blank          | 0.000 | 19:02 |
| Control Test       | 0.078 | 19:02 |
| Air Blank          | 0.000 | 19:03 |
| Diagnostics Check  | OK    | 19:03 |

---

Data Download Date/Time: 05/11/2016 10:19

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: DAWSY, B  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03415080A2  
DRY GAS EXPIRY DATE: 03/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 23:48  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: WALKER

DATE/TIME OF TEST: 04/22/2016 00:31:57

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:25 |
| Air Blank          | 0.000 | 00:25 |
| Control Test       | 0.079 | 00:25 |
| Air Blank          | 0.000 | 00:26 |
| Subject Sample # 1 | 0.103 | 00:27 |
| Breath Volume      | 2.738 | 00:27 |
| Air Blank          | 0.000 | 00:27 |
| Air Blank          | 0.000 | 00:29 |
| Subject Sample # 2 | 0.102 | 00:30 |
| Breath Volume      | 1.597 | 00:30 |
| Air Blank          | 0.000 | 00:31 |
| Control Test       | 0.077 | 00:31 |
| Air Blank          | 0.000 | 00:31 |
| Diagnostics Check  | OK    | 00:31 |

---

Data Download Date/Time: 05/11/2016 10:19

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TORRES,JENIFER,  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BRIGHTWELL,R  
ARRESTING OFFICER AGENCY: U of So FL Campus PD  
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03415080A2  
DRY GAS EXPIRY DATE: 03/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 02:48  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: STASIO

DATE/TIME OF TEST: 04/22/2016 03:54:20

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:47 |
| Air Blank          | 0.000 | 03:47 |
| Control Test       | 0.078 | 03:47 |
| Air Blank          | 0.000 | 03:48 |
| Subject Sample # 1 | 0.200 | 03:49 |
| Breath Volume      | 2.718 | 03:49 |
| Air Blank          | 0.000 | 03:50 |
| Air Blank          | 0.000 | 03:51 |
| Subject Sample # 2 | 0.192 | 03:52 |
| Breath Volume      | 1.582 | 03:52 |
| Air Blank          | 0.000 | 03:53 |
| Control Test       | 0.077 | 03:53 |
| Air Blank          | 0.000 | 03:54 |
| Diagnostics Check  | OK    | 03:54 |

---

Data Download Date/Time: 05/11/2016 10:19

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HENRY, PATRCIA,  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: WOODEN, D  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03415080A2  
DRY GAS EXPIRY DATE: 03/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 05:01  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 05:54

SUBJECT'S LAST NAME: LOHN

DATE/TIME OF TEST: 04/22/2016 06:21:04

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 06:14 |
| Air Blank          | 0.000 | 06:14 |
| Control Test       | 0.078 | 06:15 |
| Air Blank          | 0.000 | 06:15 |
| Subject Sample # 1 | 0.196 | 06:16 |
| Breath Volume      | 1.683 | 06:16 |
| Air Blank          | 0.000 | 06:17 |
| Air Blank          | 0.000 | 06:19 |
| Subject Sample # 2 | 0.213 | 06:19 |
| Breath Volume      | 1.777 | 06:19 |
| Air Blank          | 0.000 | 06:20 |
| Control Test       | 0.077 | 06:20 |
| Air Blank          | 0.000 | 06:20 |
| Diagnostics Check  | OK    | 06:21 |

---

Data Download Date/Time: 05/11/2016 10:19



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: SIMPKINS, D  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03415080A2  
DRY GAS EXPIRY DATE: 03/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 22:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:24

SUBJECT'S LAST NAME: VASQUEZ

DATE/TIME OF TEST: 04/22/2016 22:50:51

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:44 |
| Air Blank          | 0.000 | 22:45 |
| Control Test       | 0.078 | 22:45 |
| Air Blank          | 0.000 | 22:45 |
| Subject Sample # 1 | 0.176 | 22:46 |
| Breath Volume      | 2.953 | 22:46 |
| Air Blank          | 0.000 | 22:46 |
| Air Blank          | 0.000 | 22:48 |
| Subject Sample # 2 | 0.184 | 22:49 |
| Breath Volume      | 2.328 | 22:49 |
| Air Blank          | 0.000 | 22:49 |
| Control Test       | 0.078 | 22:50 |
| Air Blank          | 0.000 | 22:50 |
| Diagnostics Check  | OK    | 22:50 |

---

Data Download Date/Time: 05/11/2016 10:19

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HENRY, PATRICIA,  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: VAN-TREESE, S  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03415080A2  
DRY GAS EXPIRY DATE: 03/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 01:39  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: EPPS

DATE/TIME OF TEST: 04/23/2016 02:22:46

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:16 |
| Air Blank          | 0.000 | 02:16 |
| Control Test       | 0.079 | 02:16 |
| Air Blank          | 0.000 | 02:17 |
| Subject Sample # 1 | 0.187 | 02:18 |
| Breath Volume      | 2.394 | 02:18 |
| Air Blank          | 0.000 | 02:18 |
| Air Blank          | 0.000 | 02:20 |
| Subject Sample # 2 | 0.180 | 02:21 |
| Breath Volume      | 1.265 | 02:21 |
| Air Blank          | 0.000 | 02:21 |
| Control Test       | 0.078 | 02:22 |
| Air Blank          | 0.000 | 02:22 |
| Diagnostics Check  | OK    | 02:22 |

---

Data Download Date/Time: 05/11/2016 10:19

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TORRES,JENIFER,  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: VISSER,A  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03415080A2  
DRY GAS EXPIRY DATE: 03/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 02:05  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:34

SUBJECT'S LAST NAME: QUIROGA LOPEZ

DATE/TIME OF TEST: 04/23/2016 03:02:43

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:55 |
| Air Blank          | 0.000 | 02:56 |
| Control Test       | 0.079 | 02:56 |
| Air Blank          | 0.000 | 02:57 |
| Subject Sample # 1 | 0.039 | 02:58 |
| Breath Volume      | 1.292 | 02:58 |
| Air Blank          | 0.000 | 02:58 |
| Air Blank          | 0.000 | 03:00 |
| Subject Sample # 2 | 0.042 | 03:01 |
| Breath Volume      | 1.996 | 03:01 |
| Air Blank          | 0.000 | 03:01 |
| Control Test       | 0.079 | 03:02 |
| Air Blank          | 0.000 | 03:02 |
| Diagnostics Check  | OK    | 03:02 |

---

Data Download Date/Time: 05/11/2016 10:19

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SNYDER,MELONIE,E  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BALDWIN,C  
ARRESTING OFFICER AGENCY: Temple Terrace PD  
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03415080A2  
DRY GAS EXPIRY DATE: 03/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 06:14  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:42

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 04/23/2016 07:09:42

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 07:03 |
| Air Blank          | 0.000 | 07:03 |
| Control Test       | 0.079 | 07:03 |
| Air Blank          | 0.000 | 07:04 |
| Subject Sample # 1 | 0.126 | 07:05 |
| Breath Volume      | 1.500 | 07:05 |
| Air Blank          | 0.000 | 07:05 |
| Air Blank          | 0.000 | 07:07 |
| Subject Sample # 2 | 0.126 | 07:08 |
| Breath Volume      | 1.132 | 07:08 |
| Air Blank          | 0.000 | 07:08 |
| Control Test       | 0.078 | 07:09 |
| Air Blank          | 0.000 | 07:09 |
| Diagnostics Check  | OK    | 07:09 |

---

Data Download Date/Time: 05/11/2016 10:19

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: AGUAYO,J  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03415080A2  
DRY GAS EXPIRY DATE: 03/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 23:59  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:49

SUBJECT'S LAST NAME: SWINDLE

DATE/TIME OF TEST: 04/24/2016 01:23:18

---

|                   |           |       |                |
|-------------------|-----------|-------|----------------|
| Diagnostics Check | OK        | 01:09 |                |
| Air Blank         | 0.000     | 01:09 |                |
| Control Test      | 0.080     | 01:09 |                |
| Air Blank         | 0.000     | 01:10 |                |
| Subject Sample    | # 1 0.265 | 01:12 |                |
| Breath Volume     | 1.363     | 01:12 |                |
| Air Blank         | 0.000     | 01:13 |                |
| Air Blank         | 0.000     | 01:15 |                |
| Subject Sample    | # 2 0.245 | 01:18 | Volume Not Met |
| Breath Volume     | 0.390     | 01:18 |                |
| Air Blank         | 0.000     | 01:19 |                |
| Air Blank         | 0.000     | 01:20 |                |
| Subject Sample    | # 3 0.273 | 01:21 |                |
| Breath Volume     | 1.359     | 01:21 |                |
| Air Blank         | 0.000     | 01:22 |                |
| Control Test      | 0.075     | 01:22 |                |
| Air Blank         | 0.000     | 01:23 |                |
| Diagnostics Check | OK        | 01:23 |                |

---

Data Download Date/Time: 05/11/2016 10:19

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MCMINN,KATHY,  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: SIMPKINS,D  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03415080A2  
DRY GAS EXPIRY DATE: 03/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 01:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:22

SUBJECT'S LAST NAME: BERNAL CAICEDO

DATE/TIME OF TEST: 04/24/2016 02:55:38

---

|                   |       |       |                    |
|-------------------|-------|-------|--------------------|
| Diagnostics Check | OK    | 02:44 |                    |
| Air Blank         | 0.000 | 02:44 |                    |
| Control Test      | 0.078 | 02:44 |                    |
| Air Blank         | 0.000 | 02:45 |                    |
| Subject Sample    | # 1   | 02:48 | No Sample Provided |
| Breath Volume     | 0.000 | 02:48 |                    |
| Air Blank         | 0.000 | 02:49 |                    |
| Air Blank         | 0.000 | 02:51 |                    |
| Subject Sample    | # 2   | 02:54 | No Sample Provided |
| Breath Volume     | 0.000 | 02:54 |                    |
| Air Blank         | 0.000 | 02:54 |                    |
| Control Test      | 0.078 | 02:55 |                    |
| Air Blank         | 0.000 | 02:55 |                    |
| Diagnostics Check | OK    | 02:55 |                    |

---

Data Download Date/Time: 05/11/2016 10:19

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: AYERS,M  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03415080A2  
DRY GAS EXPIRY DATE: 03/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 02:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: BAJRACHARYA

DATE/TIME OF TEST: 04/24/2016 03:22:07

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:15 |
| Air Blank          | 0.000 | 03:16 |
| Control Test       | 0.078 | 03:16 |
| Air Blank          | 0.000 | 03:16 |
| Subject Sample # 1 | 0.184 | 03:17 |
| Breath Volume      | 2.191 | 03:17 |
| Air Blank          | 0.000 | 03:18 |
| Air Blank          | 0.000 | 03:20 |
| Subject Sample # 2 | 0.186 | 03:20 |
| Breath Volume      | 2.507 | 03:20 |
| Air Blank          | 0.000 | 03:21 |
| Control Test       | 0.077 | 03:21 |
| Air Blank          | 0.000 | 03:22 |
| Diagnostics Check  | OK    | 03:22 |

---

Data Download Date/Time: 05/11/2016 10:19

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KIRK,CASEY,S  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: ANGELICCHI,J  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03415080A2  
DRY GAS EXPIRY DATE: 03/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 03:08  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:42

SUBJECT'S LAST NAME: LEWIS-BURCH

DATE/TIME OF TEST: 04/24/2016 04:17:17

---

|                   |           |       |                |
|-------------------|-----------|-------|----------------|
| Diagnostics Check | OK        | 04:02 |                |
| Air Blank         | 0.000     | 04:03 |                |
| Control Test      | 0.079     | 04:03 |                |
| Air Blank         | 0.000     | 04:04 |                |
| Subject Sample    | # 1 0.139 | 04:07 |                |
| Breath Volume     | 1.726     | 04:07 |                |
| Air Blank         | 0.000     | 04:07 |                |
| Air Blank         | 0.000     | 04:09 |                |
| Subject Sample    | # 2 0.126 | 04:12 | Volume Not Met |
| Breath Volume     | 0.109     | 04:12 |                |
| Air Blank         | 0.000     | 04:13 |                |
| Air Blank         | 0.000     | 04:15 |                |
| Subject Sample    | # 3 0.131 | 04:15 |                |
| Breath Volume     | 1.968     | 04:15 |                |
| Air Blank         | 0.000     | 04:16 |                |
| Control Test      | 0.076     | 04:16 |                |
| Air Blank         | 0.000     | 04:17 |                |
| Diagnostics Check | OK        | 04:17 |                |

---

Data Download Date/Time: 05/11/2016 10:19



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: MANGANARO,J  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03415080A2  
DRY GAS EXPIRY DATE: 03/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 03:58  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:28

SUBJECT'S LAST NAME: CHILDS

DATE/TIME OF TEST: 04/24/2016 04:54:46

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:48 |
| Air Blank          | 0.000 | 04:48 |
| Control Test       | 0.077 | 04:49 |
| Air Blank          | 0.000 | 04:49 |
| Subject Sample # 1 | 0.104 | 04:50 |
| Breath Volume      | 2.742 | 04:50 |
| Air Blank          | 0.000 | 04:50 |
| Air Blank          | 0.000 | 04:52 |
| Subject Sample # 2 | 0.108 | 04:53 |
| Breath Volume      | 2.515 | 04:53 |
| Air Blank          | 0.000 | 04:53 |
| Control Test       | 0.078 | 04:54 |
| Air Blank          | 0.000 | 04:54 |
| Diagnostics Check  | OK    | 04:54 |

---

Data Download Date/Time: 05/11/2016 10:19

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BARNHART,H  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03415080A2  
DRY GAS EXPIRY DATE: 03/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 22:17  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:18

SUBJECT'S LAST NAME: ALCOLEA CABRERA

DATE/TIME OF TEST: 04/24/2016 23:48:25

---

|                   |           |       |                      |
|-------------------|-----------|-------|----------------------|
| Diagnostics Check | OK        | 23:39 |                      |
| Air Blank         | 0.000     | 23:39 |                      |
| Control Test      | 0.078     | 23:40 |                      |
| Air Blank         | 0.000     | 23:40 |                      |
| Subject Sample    | # 1 0.049 | 23:43 | Volume Not Met       |
| Breath Volume     | 0.015     | 23:43 |                      |
| Air Blank         | 0.000     | 23:44 |                      |
| Air Blank         | 0.000     | 23:46 |                      |
| Subject Sample    | # 2       | 23:47 | Subject Test Refused |
| Breath Volume     | 0.000     | 23:47 |                      |
| Air Blank         | 0.000     | 23:47 |                      |
| Control Test      | 0.078     | 23:47 |                      |
| Air Blank         | 0.000     | 23:48 |                      |
| Diagnostics Check | OK        | 23:48 |                      |

---

Data Download Date/Time: 05/11/2016 10:19

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BARNHART,H  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03415080A2  
DRY GAS EXPIRY DATE: 03/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 22:17  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:18

SUBJECT'S LAST NAME: ALCOLEA CABRERA

DATE/TIME OF TEST: 04/25/2016 00:04:16

---

|                   |           |       |                |
|-------------------|-----------|-------|----------------|
| Diagnostics Check | OK        | 23:52 |                |
| Air Blank         | 0.000     | 23:53 |                |
| Control Test      | 0.078     | 23:53 |                |
| Air Blank         | 0.000     | 23:54 |                |
| Subject Sample    | # 2 0.066 | 23:57 | Volume Not Met |
| Breath Volume     | 0.050     | 23:57 |                |
| Air Blank         | 0.000     | 23:57 |                |
| Air Blank         | 0.000     | 23:59 |                |
| Subject Sample    | # 1 0.073 | 00:02 | Volume Not Met |
| Breath Volume     | 0.019     | 00:02 |                |
| Air Blank         | 0.000     | 00:03 |                |
| Control Test      | 0.078     | 00:03 |                |
| Air Blank         | 0.000     | 00:04 |                |
| Diagnostics Check | OK        | 00:04 |                |

---

Data Download Date/Time: 05/11/2016 10:19

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KIRK,CASEY,S  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: EHRHART,M  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03415080A2  
DRY GAS EXPIRY DATE: 03/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 23:14  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:58

SUBJECT'S LAST NAME: SWAILS II

DATE/TIME OF TEST: 04/26/2016 00:26:52

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:20 |
| Air Blank          | 0.000 | 00:20 |
| Control Test       | 0.078 | 00:21 |
| Air Blank          | 0.000 | 00:21 |
| Subject Sample # 1 | 0.123 | 00:22 |
| Breath Volume      | 1.417 | 00:22 |
| Air Blank          | 0.000 | 00:22 |
| Air Blank          | 0.000 | 00:24 |
| Subject Sample # 2 | 0.142 | 00:25 |
| Breath Volume      | 1.777 | 00:25 |
| Air Blank          | 0.000 | 00:25 |
| Control Test       | 0.078 | 00:26 |
| Air Blank          | 0.000 | 00:26 |
| Diagnostics Check  | OK    | 00:26 |

---

Data Download Date/Time: 05/11/2016 10:19

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BARLAUG,G  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03415080A2  
DRY GAS EXPIRY DATE: 03/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 01:24  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:52

SUBJECT'S LAST NAME: VERBANAS

DATE/TIME OF TEST: 04/26/2016 02:25:55

---

|                   |           |       |                   |
|-------------------|-----------|-------|-------------------|
| Diagnostics Check | OK        | 02:12 |                   |
| Air Blank         | 0.000     | 02:12 |                   |
| Control Test      | 0.079     | 02:13 |                   |
| Air Blank         | 0.000     | 02:13 |                   |
| Subject Sample    | # 1 0.147 | 02:14 | No .020 agreement |
| Breath Volume     | 1.519     | 02:14 |                   |
| Air Blank         | 0.000     | 02:15 |                   |
| Air Blank         | 0.000     | 02:17 |                   |
| Subject Sample    | # 2 0.089 | 02:20 | Volume Not Met    |
| Breath Volume     | 0.316     | 02:20 |                   |
| Air Blank         | 0.000     | 02:21 |                   |
| Air Blank         | 0.000     | 02:22 |                   |
| Subject Sample    | # 3 0.125 | 02:24 | No .020 agreement |
| Breath Volume     | 1.933     | 02:24 |                   |
| Air Blank         | 0.000     | 02:24 |                   |
| Control Test      | 0.077     | 02:25 |                   |
| Air Blank         | 0.000     | 02:25 |                   |
| Diagnostics Check | OK        | 02:25 |                   |

---

Data Download Date/Time: 05/11/2016 10:19

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BARLAUG,G  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03415080A2  
DRY GAS EXPIRY DATE: 03/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 01:24  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:52

SUBJECT'S LAST NAME: VERBANAS

DATE/TIME OF TEST: 04/26/2016 02:40:04

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:29 |
| Air Blank          | 0.000 | 02:29 |
| Control Test       | 0.078 | 02:30 |
| Air Blank          | 0.000 | 02:30 |
| Subject Sample # 1 | 0.150 | 02:32 |
| Breath Volume      | 2.652 | 02:32 |
| Air Blank          | 0.000 | 02:33 |
| Air Blank          | 0.000 | 02:34 |
| Subject Sample # 2 | 0.115 | 02:35 |
| Breath Volume      | 1.347 | 02:35 |
| Air Blank          | 0.000 | 02:36 |
| Air Blank          | 0.000 | 02:37 |
| Subject Sample # 3 | 0.143 | 02:38 |
| Breath Volume      | 2.316 | 02:38 |
| Air Blank          | 0.000 | 02:39 |
| Control Test       | 0.077 | 02:39 |
| Air Blank          | 0.000 | 02:39 |
| Diagnostics Check  | OK    | 02:40 |

---

Data Download Date/Time: 05/11/2016 10:19

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MCMINN,KATHY,  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: VAN TREESE,S  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03415080A2  
DRY GAS EXPIRY DATE: 03/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 03:28  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:57

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 04/26/2016 04:24:41

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:17 |
| Air Blank          | 0.000 | 04:17 |
| Control Test       | 0.078 | 04:18 |
| Air Blank          | 0.000 | 04:18 |
| Subject Sample # 1 | 0.110 | 04:20 |
| Breath Volume      | 3.347 | 04:20 |
| Air Blank          | 0.000 | 04:20 |
| Air Blank          | 0.000 | 04:22 |
| Subject Sample # 2 | 0.099 | 04:23 |
| Breath Volume      | 1.105 | 04:23 |
| Air Blank          | 0.000 | 04:23 |
| Control Test       | 0.077 | 04:24 |
| Air Blank          | 0.000 | 04:24 |
| Diagnostics Check  | OK    | 04:24 |

---

Data Download Date/Time: 05/11/2016 10:19

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: LYON, M  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03415080A2  
DRY GAS EXPIRY DATE: 03/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 21:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:54

SUBJECT'S LAST NAME: GANTER III

DATE/TIME OF TEST: 04/26/2016 22:26:13

---

|                   |           |       |                    |
|-------------------|-----------|-------|--------------------|
| Diagnostics Check | OK        | 22:14 |                    |
| Air Blank         | 0.000     | 22:15 |                    |
| Control Test      | 0.079     | 22:15 |                    |
| Air Blank         | 0.000     | 22:16 |                    |
| Subject Sample    | # 1 0.179 | 22:19 | Volume Not Met     |
| Breath Volume     | 0.312     | 22:19 |                    |
| Air Blank         | 0.000     | 22:19 |                    |
| Air Blank         | 0.000     | 22:21 |                    |
| Subject Sample    | # 2       | 22:24 | No Sample Provided |
| Breath Volume     | 0.000     | 22:24 |                    |
| Air Blank         | 0.000     | 22:25 |                    |
| Control Test      | 0.077     | 22:25 |                    |
| Air Blank         | 0.000     | 22:26 |                    |
| Diagnostics Check | OK        | 22:26 |                    |

---

Data Download Date/Time: 05/11/2016 10:19



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TORRES,JENIFER,  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BAILEY,D  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03415080A2  
DRY GAS EXPIRY DATE: 03/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 00:58  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:24

SUBJECT'S LAST NAME: GORSKI

DATE/TIME OF TEST: 04/27/2016 01:54:50

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:45 |
| Air Blank          | 0.000 | 01:46 |
| Control Test       | 0.078 | 01:46 |
| Air Blank          | 0.000 | 01:47 |
| Subject Sample # 1 | 0.147 | 01:49 |
| Breath Volume      | 1.601 | 01:49 |
| Air Blank          | 0.000 | 01:50 |
| Air Blank          | 0.000 | 01:52 |
| Subject Sample # 2 | 0.150 | 01:53 |
| Breath Volume      | 1.718 | 01:53 |
| Air Blank          | 0.000 | 01:53 |
| Control Test       | 0.077 | 01:54 |
| Air Blank          | 0.000 | 01:54 |
| Diagnostics Check  | OK    | 01:54 |

---

Data Download Date/Time: 05/11/2016 10:19

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HENRY,PATRICIA,  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: AGUAYO,J  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03415080A2  
DRY GAS EXPIRY DATE: 03/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 18:40  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 19:30

SUBJECT'S LAST NAME: LAWSON

DATE/TIME OF TEST: 04/27/2016 19:57:06

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 19:50 |
| Air Blank          | 0.000 | 19:51 |
| Control Test       | 0.079 | 19:51 |
| Air Blank          | 0.000 | 19:51 |
| Subject Sample # 1 | 0.000 | 19:52 |
| Breath Volume      | 2.394 | 19:52 |
| Air Blank          | 0.000 | 19:53 |
| Air Blank          | 0.000 | 19:55 |
| Subject Sample # 2 | 0.000 | 19:55 |
| Breath Volume      | 2.546 | 19:55 |
| Air Blank          | 0.000 | 19:56 |
| Control Test       | 0.079 | 19:56 |
| Air Blank          | 0.000 | 19:57 |
| Diagnostics Check  | OK    | 19:57 |

---

Data Download Date/Time: 05/11/2016 10:19

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KIRK,CASEY,S  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: ANDERSON,C  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03415080A2  
DRY GAS EXPIRY DATE: 03/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 02:37  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:13

SUBJECT'S LAST NAME: HARRELL

DATE/TIME OF TEST: 04/29/2016 03:42:03

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:34 |
| Air Blank          | 0.000 | 03:35 |
| Control Test       | 0.079 | 03:35 |
| Air Blank          | 0.000 | 03:36 |
| Subject Sample # 1 | 0.141 | 03:37 |
| Breath Volume      | 3.000 | 03:37 |
| Air Blank          | 0.000 | 03:38 |
| Air Blank          | 0.000 | 03:40 |
| Subject Sample # 2 | 0.138 | 03:40 |
| Breath Volume      | 3.011 | 03:40 |
| Air Blank          | 0.000 | 03:41 |
| Control Test       | 0.077 | 03:41 |
| Air Blank          | 0.000 | 03:41 |
| Diagnostics Check  | OK    | 03:42 |

---

Data Download Date/Time: 05/11/2016 10:20

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SMITH,REBECCA,J  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: LARA,A  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03415080A2  
DRY GAS EXPIRY DATE: 03/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 06:01  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:35

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 04/29/2016 07:05:33

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 06:58 |
| Air Blank          | 0.000 | 06:58 |
| Control Test       | 0.078 | 06:59 |
| Air Blank          | 0.000 | 06:59 |
| Subject Sample # 1 | 0.155 | 07:00 |
| Breath Volume      | 2.234 | 07:00 |
| Air Blank          | 0.000 | 07:01 |
| Air Blank          | 0.000 | 07:03 |
| Subject Sample # 2 | 0.152 | 07:03 |
| Breath Volume      | 2.464 | 07:03 |
| Air Blank          | 0.000 | 07:04 |
| Control Test       | 0.077 | 07:04 |
| Air Blank          | 0.000 | 07:05 |
| Diagnostics Check  | OK    | 07:05 |

---

Data Download Date/Time: 05/11/2016 10:20

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SKIPPER,ROGER,G  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BATEMAN,R 4704  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03415080A2  
DRY GAS EXPIRY DATE: 03/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 09:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 10:20

SUBJECT'S LAST NAME: MURPHY

DATE/TIME OF TEST: 04/29/2016 10:49:42

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 10:43 |
| Air Blank          | 0.000 | 10:44 |
| Control Test       | 0.078 | 10:44 |
| Air Blank          | 0.000 | 10:44 |
| Subject Sample # 1 | 0.205 | 10:45 |
| Breath Volume      | 2.722 | 10:45 |
| Air Blank          | 0.000 | 10:46 |
| Air Blank          | 0.000 | 10:47 |
| Subject Sample # 2 | 0.193 | 10:48 |
| Breath Volume      | 1.511 | 10:48 |
| Air Blank          | 0.000 | 10:48 |
| Control Test       | 0.078 | 10:49 |
| Air Blank          | 0.000 | 10:49 |
| Diagnostics Check  | OK    | 10:49 |

---

Data Download Date/Time: 05/11/2016 10:20

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MCMINN,KATHY,  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: SNIDER,P  
ARRESTING OFFICER AGENCY: Plant City PD  
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03415080A2  
DRY GAS EXPIRY DATE: 03/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 20:33  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:27

SUBJECT'S LAST NAME: GARCIA-MENDEZ

DATE/TIME OF TEST: 04/29/2016 21:55:39

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 21:48 |
| Air Blank          | 0.000 | 21:48 |
| Control Test       | 0.078 | 21:48 |
| Air Blank          | 0.000 | 21:49 |
| Subject Sample # 1 | 0.180 | 21:50 |
| Breath Volume      | 1.523 | 21:50 |
| Air Blank          | 0.000 | 21:51 |
| Air Blank          | 0.000 | 21:52 |
| Subject Sample # 2 | 0.181 | 21:54 |
| Breath Volume      | 1.585 | 21:54 |
| Air Blank          | 0.000 | 21:54 |
| Control Test       | 0.077 | 21:55 |
| Air Blank          | 0.000 | 21:55 |
| Diagnostics Check  | OK    | 21:55 |

---

Data Download Date/Time: 05/11/2016 10:20

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KIRK,CASEY,S  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: FRANKLIN,R  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03415080A2  
DRY GAS EXPIRY DATE: 03/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 22:46  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: BAROWSKI

DATE/TIME OF TEST: 04/29/2016 23:46:53

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:40 |
| Air Blank          | 0.000 | 23:41 |
| Control Test       | 0.078 | 23:41 |
| Air Blank          | 0.000 | 23:42 |
| Subject Sample # 1 | 0.078 | 23:42 |
| Breath Volume      | 2.714 | 23:42 |
| Air Blank          | 0.000 | 23:43 |
| Air Blank          | 0.000 | 23:44 |
| Subject Sample # 2 | 0.078 | 23:45 |
| Breath Volume      | 2.316 | 23:45 |
| Air Blank          | 0.000 | 23:45 |
| Control Test       | 0.078 | 23:46 |
| Air Blank          | 0.000 | 23:46 |
| Diagnostics Check  | OK    | 23:46 |

---

Data Download Date/Time: 05/11/2016 10:20

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MCMINN,KATHY,  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: SNIDER,P  
ARRESTING OFFICER AGENCY: Plant City PD  
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03415080A2  
DRY GAS EXPIRY DATE: 03/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 01:23  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:17

SUBJECT'S LAST NAME: HORN

DATE/TIME OF TEST: 04/30/2016 02:46:29

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:39 |
| Air Blank          | 0.000 | 02:40 |
| Control Test       | 0.079 | 02:40 |
| Air Blank          | 0.000 | 02:41 |
| Subject Sample # 1 | 0.177 | 02:42 |
| Breath Volume      | 1.382 | 02:42 |
| Air Blank          | 0.000 | 02:42 |
| Air Blank          | 0.000 | 02:44 |
| Subject Sample # 2 | 0.178 | 02:44 |
| Breath Volume      | 1.296 | 02:44 |
| Air Blank          | 0.000 | 02:45 |
| Control Test       | 0.078 | 02:45 |
| Air Blank          | 0.000 | 02:46 |
| Diagnostics Check  | OK    | 02:46 |

---

Data Download Date/Time: 05/11/2016 10:20



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MCMINN,KATHY,  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BRENNAN,D  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03415080A2  
DRY GAS EXPIRY DATE: 03/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 03:51  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:17

SUBJECT'S LAST NAME: NARVAEZ

DATE/TIME OF TEST: 04/30/2016 04:44:05

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:37 |
| Air Blank          | 0.000 | 04:38 |
| Control Test       | 0.079 | 04:38 |
| Air Blank          | 0.000 | 04:39 |
| Subject Sample # 1 | 0.144 | 04:39 |
| Breath Volume      | 3.015 | 04:39 |
| Air Blank          | 0.000 | 04:40 |
| Air Blank          | 0.000 | 04:42 |
| Subject Sample # 2 | 0.141 | 04:42 |
| Breath Volume      | 2.000 | 04:42 |
| Air Blank          | 0.000 | 04:43 |
| Control Test       | 0.079 | 04:43 |
| Air Blank          | 0.000 | 04:44 |
| Diagnostics Check  | OK    | 04:44 |

---

Data Download Date/Time: 05/11/2016 10:20

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARR,BARBARA,H  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: THOMPSON,B  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03415080A2  
DRY GAS EXPIRY DATE: 03/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 17:47  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:17

SUBJECT'S LAST NAME: REINER

DATE/TIME OF TEST: 04/30/2016 18:43:40

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 18:37 |
| Air Blank          | 0.000 | 18:38 |
| Control Test       | 0.079 | 18:38 |
| Air Blank          | 0.000 | 18:38 |
| Subject Sample # 1 | 0.000 | 18:39 |
| Breath Volume      | 3.367 | 18:39 |
| Air Blank          | 0.000 | 18:39 |
| Air Blank          | 0.000 | 18:41 |
| Subject Sample # 2 | 0.000 | 18:42 |
| Breath Volume      | 2.921 | 18:42 |
| Air Blank          | 0.000 | 18:42 |
| Control Test       | 0.079 | 18:43 |
| Air Blank          | 0.000 | 18:43 |
| Diagnostics Check  | OK    | 18:43 |

---

Data Download Date/Time: 05/11/2016 10:20

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GAGO-  
CHUQUILLANQUI,ALBERTO,E  
BREATH TEST OPERATOR AGENCY: Marion County SO  
ARRESTING OFFICER: GAGO-  
CHUQUILLANQUI,ALBERTO  
ARRESTING OFFICER AGENCY: Marion County SO  
LAST AGENCY INSPECTION: 03/24/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 83550  
DRY GAS EXPIRY DATE: 09/26/2018

VIOLATION CODE: DUI  
ARREST TIME: 23:13  
GENDER: Female

OBSERVATION PERIOD BEGAN: 23:54

SUBJECT'S LAST NAME: RAMSEY

DATE/TIME OF TEST: 04/02/2016 00:20:52

---

|                   |           |       |            |
|-------------------|-----------|-------|------------|
| Diagnostics Check | OK        | 00:16 |            |
| Air Blank         | 0.000     | 00:16 |            |
| Control Test      | 0.080     | 00:17 |            |
| Air Blank         | 0.000     | 00:17 |            |
| Subject Sample    | # 1 0.133 | 00:19 |            |
| Breath Volume     | 1.566     | 00:19 |            |
| Air Blank         |           | 00:20 | Purge Fail |
| Air Blank         | 0.000     | 00:20 |            |

---

Data Download Date/Time: 04/28/2016 10:38

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GAGO-  
CHUQUILLANQUI,ALBERTO,E  
BREATH TEST OPERATOR AGENCY: Marion County SO  
ARRESTING OFFICER: GAGO-  
CHUQUILLANQUI,ALBERTO  
ARRESTING OFFICER AGENCY: Marion County SO  
LAST AGENCY INSPECTION: 03/24/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 83550  
DRY GAS EXPIRY DATE: 09/26/2018

VIOLATION CODE: DUI  
ARREST TIME: 23:13  
GENDER: Female

OBSERVATION PERIOD BEGAN: 23:54

SUBJECT'S LAST NAME: RAMSEY

DATE/TIME OF TEST: 04/02/2016 00:27:11

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 00:22 |                      |
| Air Blank         | 0.000 | 00:23 |                      |
| Control Test      | 0.078 | 00:23 |                      |
| Air Blank         | 0.000 | 00:24 |                      |
| Subject Sample    | # 1   | 00:25 | Subject Test Refused |
| Breath Volume     | 0.000 | 00:25 |                      |
| Air Blank         | 0.000 | 00:26 |                      |
| Control Test      | 0.079 | 00:26 |                      |
| Air Blank         | 0.000 | 00:27 |                      |
| Diagnostics Check | OK    | 00:27 |                      |

---

Data Download Date/Time: 04/28/2016 10:38

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SOUZA,MAYKOLL,G  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: POWERS,CHRISTOPHER  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 03/24/2016

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 83550  
DRY GAS EXPIRY DATE: 09/26/2018  
VIOLATION CODE: DUI  
ARREST TIME: 21:11  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:51

SUBJECT'S LAST NAME: YOSKIN

DATE/TIME OF TEST: 04/03/2016 22:28:06

---

|                   |           |       |                |
|-------------------|-----------|-------|----------------|
| Diagnostics Check | OK        | 22:11 |                |
| Air Blank         | 0.000     | 22:11 |                |
| Control Test      | 0.080     | 22:12 |                |
| Air Blank         | 0.000     | 22:12 |                |
| Subject Sample    | # 1 0.124 | 22:15 | Volume Not Met |
| Breath Volume     | 0.257     | 22:15 |                |
| Air Blank         | 0.000     | 22:16 |                |
| Air Blank         | 0.000     | 22:18 |                |
| Subject Sample    | # 2 0.128 | 22:21 |                |
| Breath Volume     | 1.507     | 22:21 |                |
| Air Blank         | 0.000     | 22:22 |                |
| Air Blank         | 0.000     | 22:24 |                |
| Subject Sample    | # 3 0.124 | 22:26 |                |
| Breath Volume     | 1.683     | 22:26 |                |
| Air Blank         | 0.000     | 22:27 |                |
| Control Test      | 0.078     | 22:27 |                |
| Air Blank         | 0.000     | 22:28 |                |
| Diagnostics Check | OK        | 22:28 |                |

---

Data Download Date/Time: 04/28/2016 10:38

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DOBOSIEWICZ, ANTHONY, N  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: DOBOSIEWICZ, ANTHONY  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 03/24/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 83550  
DRY GAS EXPIRY DATE: 09/26/2018  
VIOLATION CODE: DUI  
ARREST TIME: 22:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: FELIX OBESO

DATE/TIME OF TEST: 04/04/2016 23:17:54

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:11 |
| Air Blank          | 0.000 | 23:11 |
| Control Test       | 0.079 | 23:12 |
| Air Blank          | 0.000 | 23:12 |
| Subject Sample # 1 | 0.152 | 23:13 |
| Breath Volume      | 1.550 | 23:13 |
| Air Blank          | 0.000 | 23:13 |
| Air Blank          | 0.000 | 23:15 |
| Subject Sample # 2 | 0.163 | 23:16 |
| Breath Volume      | 2.851 | 23:16 |
| Air Blank          | 0.000 | 23:16 |
| Control Test       | 0.081 | 23:17 |
| Air Blank          | 0.000 | 23:17 |
| Diagnostics Check  | OK    | 23:17 |

---

Data Download Date/Time: 04/28/2016 10:38

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PUBILLONES,GEOFFREY,A  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: PUBILLONES,GEOFFREY  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 03/24/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1662060  
DRY GAS EXPIRY DATE: 04/18/2017  
VIOLATION CODE: DUI  
ARREST TIME: 01:00  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: MOORE

DATE/TIME OF TEST: 04/05/2016 01:40:12

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 01:37 |                      |
| Air Blank         | 0.000 | 01:37 |                      |
| Control Test      | 0.079 | 01:38 |                      |
| Air Blank         | 0.000 | 01:38 |                      |
| Subject Sample    | # 1   | 01:38 | Subject Test Refused |
| Breath Volume     | 0.000 | 01:38 |                      |
| Air Blank         | 0.000 | 01:39 |                      |
| Control Test      | 0.080 | 01:39 |                      |
| Air Blank         | 0.000 | 01:40 |                      |
| Diagnostics Check | OK    | 01:40 |                      |

---

Data Download Date/Time: 04/28/2016 10:38

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MANGUM,RACHEL,E  
BREATH TEST OPERATOR AGENCY: Ocala PD  
ARRESTING OFFICER: SIGLER,SHAWN  
ARRESTING OFFICER AGENCY: Ocala PD  
LAST AGENCY INSPECTION: 03/24/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1662060  
DRY GAS EXPIRY DATE: 04/18/2017  
VIOLATION CODE: DUI  
ARREST TIME: 04:08  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:31

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 04/08/2016 04:58:32

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:52 |
| Air Blank          | 0.000 | 04:52 |
| Control Test       | 0.079 | 04:52 |
| Air Blank          | 0.000 | 04:53 |
| Subject Sample # 1 | 0.190 | 04:53 |
| Breath Volume      | 2.632 | 04:53 |
| Air Blank          | 0.000 | 04:54 |
| Air Blank          | 0.000 | 04:56 |
| Subject Sample # 2 | 0.197 | 04:56 |
| Breath Volume      | 1.675 | 04:56 |
| Air Blank          | 0.000 | 04:57 |
| Control Test       | 0.081 | 04:57 |
| Air Blank          | 0.000 | 04:58 |
| Diagnostics Check  | OK    | 04:58 |

---

Data Download Date/Time: 04/28/2016 10:38



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DOBOSIEWICZ,ANTHONY,N  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: DOBOSIEWICZ,ANTHONY  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 03/24/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 83550  
DRY GAS EXPIRY DATE: 09/26/2018  
VIOLATION CODE: DUI  
ARREST TIME: 16:47  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:00

SUBJECT'S LAST NAME: ACOSTA RIVERA

DATE/TIME OF TEST: 04/08/2016 17:54:07

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 17:47 |
| Air Blank          | 0.000 | 17:48 |
| Control Test       | 0.080 | 17:48 |
| Air Blank          | 0.000 | 17:48 |
| Subject Sample # 1 | 0.205 | 17:49 |
| Breath Volume      | 1.246 | 17:49 |
| Air Blank          | 0.000 | 17:50 |
| Air Blank          | 0.000 | 17:51 |
| Subject Sample # 2 | 0.192 | 17:52 |
| Breath Volume      | 1.902 | 17:52 |
| Air Blank          | 0.000 | 17:53 |
| Control Test       | 0.080 | 17:53 |
| Air Blank          | 0.000 | 17:54 |
| Diagnostics Check  | OK    | 17:54 |

---

Data Download Date/Time: 04/28/2016 10:38

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MILLER,GARY,A  
BREATH TEST OPERATOR AGENCY: Marion County SO  
ARRESTING OFFICER: MILLER,GARY  
ARRESTING OFFICER AGENCY: Marion County SO  
LAST AGENCY INSPECTION: 03/24/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 83550  
DRY GAS EXPIRY DATE: 09/26/2018  
VIOLATION CODE: DUI  
ARREST TIME: 02:31  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:53

SUBJECT'S LAST NAME: HADD

DATE/TIME OF TEST: 04/09/2016 03:19:46

---

|                   |       |       |                    |
|-------------------|-------|-------|--------------------|
| Diagnostics Check | OK    | 03:13 |                    |
| Air Blank         | 0.000 | 03:14 |                    |
| Control Test      | 0.079 | 03:14 |                    |
| Air Blank         | 0.000 | 03:15 |                    |
| Subject Sample    | # 1   | 03:17 | Interferent Detect |
| Breath Volume     | 1.871 | 03:17 |                    |
| Air Blank         | 0.000 | 03:18 |                    |
| Control Test      | 0.082 | 03:19 |                    |
| Air Blank         | 0.000 | 03:19 |                    |
| Diagnostics Check | OK    | 03:19 |                    |

---

Data Download Date/Time: 04/28/2016 10:38

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MILLER,GARY,A  
BREATH TEST OPERATOR AGENCY: Marion County SO  
ARRESTING OFFICER: MILLER,GARY  
ARRESTING OFFICER AGENCY: Marion County SO  
LAST AGENCY INSPECTION: 03/24/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 83550  
DRY GAS EXPIRY DATE: 09/26/2018  
VIOLATION CODE: DUI  
ARREST TIME: 02:31  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:53

SUBJECT'S LAST NAME: HADD

DATE/TIME OF TEST: 04/09/2016 03:28:59

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:21 |
| Air Blank          | 0.000 | 03:22 |
| Control Test       | 0.079 | 03:22 |
| Air Blank          | 0.000 | 03:22 |
| Subject Sample # 1 | 0.121 | 03:23 |
| Breath Volume      | 1.605 | 03:23 |
| Air Blank          | 0.000 | 03:23 |
| Air Blank          | 0.000 | 03:25 |
| Subject Sample # 2 | 0.122 | 03:27 |
| Breath Volume      | 2.136 | 03:27 |
| Air Blank          | 0.000 | 03:28 |
| Control Test       | 0.080 | 03:28 |
| Air Blank          | 0.000 | 03:28 |
| Diagnostics Check  | OK    | 03:28 |

---

Data Download Date/Time: 04/28/2016 10:38

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LETSON,REBECCA,G  
BREATH TEST OPERATOR AGENCY: Ocala PD  
ARRESTING OFFICER: LETSON,REBECCA  
ARRESTING OFFICER AGENCY: Ocala PD  
LAST AGENCY INSPECTION: 03/24/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 83550  
DRY GAS EXPIRY DATE: 09/26/2018  
VIOLATION CODE: DUI  
ARREST TIME: 23:26  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: ROBERTS JR

DATE/TIME OF TEST: 04/10/2016 00:22:56

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:15 |
| Air Blank          | 0.000 | 00:16 |
| Control Test       | 0.077 | 00:16 |
| Air Blank          | 0.000 | 00:17 |
| Subject Sample # 1 | 0.158 | 00:18 |
| Breath Volume      | 4.195 | 00:18 |
| Air Blank          | 0.000 | 00:18 |
| Air Blank          | 0.000 | 00:20 |
| Subject Sample # 2 | 0.158 | 00:21 |
| Breath Volume      | 4.062 | 00:21 |
| Air Blank          | 0.000 | 00:22 |
| Control Test       | 0.081 | 00:22 |
| Air Blank          | 0.000 | 00:22 |
| Diagnostics Check  | OK    | 00:22 |

---

Data Download Date/Time: 04/28/2016 10:38

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WINKLER,TODD,A  
BREATH TEST OPERATOR AGENCY: Marion County SO  
ARRESTING OFFICER: LOCKHART,JESSE  
ARRESTING OFFICER AGENCY: Marion County SO  
LAST AGENCY INSPECTION: 03/24/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 83550  
DRY GAS EXPIRY DATE: 09/26/2018  
VIOLATION CODE: DUI  
ARREST TIME: 15:06  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 15:55

SUBJECT'S LAST NAME: HOLLISTER

DATE/TIME OF TEST: 04/10/2016 16:21:32

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 16:15 |
| Air Blank          | 0.000 | 16:16 |
| Control Test       | 0.080 | 16:16 |
| Air Blank          | 0.000 | 16:16 |
| Subject Sample # 1 | 0.000 | 16:17 |
| Breath Volume      | 2.148 | 16:17 |
| Air Blank          | 0.000 | 16:17 |
| Air Blank          | 0.000 | 16:19 |
| Subject Sample # 2 | 0.000 | 16:20 |
| Breath Volume      | 1.382 | 16:20 |
| Air Blank          | 0.000 | 16:20 |
| Control Test       | 0.080 | 16:20 |
| Air Blank          | 0.000 | 16:21 |
| Diagnostics Check  | OK    | 16:21 |

---

Data Download Date/Time: 04/28/2016 10:38

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FERNANDEZ,ALAIN,  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: FERNANDEZ,ALAIN  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 03/24/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 83550  
DRY GAS EXPIRY DATE: 09/26/2018  
VIOLATION CODE: DUI  
ARREST TIME: 13:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 14:19

SUBJECT'S LAST NAME: BUTLER

DATE/TIME OF TEST: 04/11/2016 14:42:29

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 14:39 |                      |
| Air Blank         | 0.000 | 14:40 |                      |
| Control Test      | 0.080 | 14:40 |                      |
| Air Blank         | 0.000 | 14:40 |                      |
| Subject Sample    | # 1   | 14:41 | Subject Test Refused |
| Breath Volume     | 0.000 | 14:41 |                      |
| Air Blank         | 0.000 | 14:41 |                      |
| Control Test      | 0.080 | 14:41 |                      |
| Air Blank         | 0.000 | 14:42 |                      |
| Diagnostics Check | OK    | 14:42 |                      |

---

Data Download Date/Time: 04/28/2016 10:38

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CARPENTER,DONALD,M  
BREATH TEST OPERATOR AGENCY: Marion County SO  
ARRESTING OFFICER: MASLAN,JESSE  
ARRESTING OFFICER AGENCY: Marion County SO  
LAST AGENCY INSPECTION: 03/24/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 83550  
DRY GAS EXPIRY DATE: 09/26/2018  
VIOLATION CODE: DUI  
ARREST TIME: 23:53  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:19

SUBJECT'S LAST NAME: CARBO

DATE/TIME OF TEST: 04/12/2016 00:43:03

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 00:40 |                      |
| Air Blank         | 0.000 | 00:40 |                      |
| Control Test      | 0.080 | 00:40 |                      |
| Air Blank         | 0.000 | 00:41 |                      |
| Subject Sample    | # 1   | 00:41 | Subject Test Refused |
| Breath Volume     | 0.000 | 00:41 |                      |
| Air Blank         | 0.000 | 00:42 |                      |
| Control Test      | 0.079 | 00:42 |                      |
| Air Blank         | 0.000 | 00:42 |                      |
| Diagnostics Check | OK    | 00:43 |                      |

---

Data Download Date/Time: 04/28/2016 10:38

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WHITMORE,JAMES,D  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: WHIMTORE,JAMES  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 03/24/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1662060  
DRY GAS EXPIRY DATE: 04/18/2017  
VIOLATION CODE: DUI  
ARREST TIME: 19:45  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 19:30

SUBJECT'S LAST NAME: REISS

DATE/TIME OF TEST: 04/15/2016 20:11:36

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 20:04 |
| Air Blank          | 0.000 | 20:05 |
| Control Test       | 0.079 | 20:05 |
| Air Blank          | 0.000 | 20:05 |
| Subject Sample # 1 | 0.192 | 20:06 |
| Breath Volume      | 1.859 | 20:06 |
| Air Blank          | 0.000 | 20:07 |
| Air Blank          | 0.000 | 20:08 |
| Subject Sample # 2 | 0.182 | 20:10 |
| Breath Volume      | 1.488 | 20:10 |
| Air Blank          | 0.000 | 20:10 |
| Control Test       | 0.079 | 20:10 |
| Air Blank          | 0.000 | 20:11 |
| Diagnostics Check  | OK    | 20:11 |

---

Data Download Date/Time: 04/28/2016 10:38



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DELRUSSO,CHRISTOPHER,M  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: DELRUSSO,CHRISTOPHER  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 03/24/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 83550  
DRY GAS EXPIRY DATE: 09/26/2018  
VIOLATION CODE: DUI  
ARREST TIME: 23:38  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:06

SUBJECT'S LAST NAME: ROBERTSON

DATE/TIME OF TEST: 04/17/2016 00:41:30

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:35 |
| Air Blank          | 0.000 | 00:35 |
| Control Test       | 0.080 | 00:35 |
| Air Blank          | 0.000 | 00:36 |
| Subject Sample # 1 | 0.289 | 00:36 |
| Breath Volume      | 2.355 | 00:36 |
| Air Blank          | 0.000 | 00:37 |
| Air Blank          | 0.000 | 00:39 |
| Subject Sample # 2 | 0.288 | 00:39 |
| Breath Volume      | 2.589 | 00:39 |
| Air Blank          | 0.000 | 00:40 |
| Control Test       | 0.080 | 00:40 |
| Air Blank          | 0.000 | 00:41 |
| Diagnostics Check  | OK    | 00:41 |

---

Data Download Date/Time: 04/28/2016 10:38

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PUBILLONES,GEOFFREY,A  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: PUBILLONES,GEOFFREY  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 03/24/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1662060  
DRY GAS EXPIRY DATE: 04/18/2017  
VIOLATION CODE: DUI  
ARREST TIME: 22:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: CARNEGIA

DATE/TIME OF TEST: 04/19/2016 00:16:32

---

|                   |           |       |                |
|-------------------|-----------|-------|----------------|
| Diagnostics Check | OK        | 00:02 |                |
| Air Blank         | 0.000     | 00:02 |                |
| Control Test      | 0.079     | 00:02 |                |
| Air Blank         | 0.000     | 00:03 |                |
| Subject Sample    | # 1 0.123 | 00:06 | Volume Not Met |
| Breath Volume     | 0.000     | 00:06 |                |
| Air Blank         | 0.000     | 00:07 |                |
| Air Blank         | 0.000     | 00:08 |                |
| Subject Sample    | # 2 0.127 | 00:09 |                |
| Breath Volume     | 1.582     | 00:09 |                |
| Air Blank         | 0.000     | 00:10 |                |
| Air Blank         | 0.000     | 00:12 |                |
| Subject Sample    | # 3 0.126 | 00:15 | Volume Not Met |
| Breath Volume     | 0.214     | 00:15 |                |
| Air Blank         |           | 00:16 | Purge Fail     |
| Air Blank         | 0.000     | 00:16 |                |

---

Data Download Date/Time: 04/28/2016 10:38

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GALLER,JESSICA,G  
BREATH TEST OPERATOR AGENCY: Marion County SO  
ARRESTING OFFICER: GALLER,CHRISTOPHER  
ARRESTING OFFICER AGENCY: Marion County SO  
LAST AGENCY INSPECTION: 03/24/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 83550  
DRY GAS EXPIRY DATE: 09/26/2018  
VIOLATION CODE: DUI  
ARREST TIME: 00:11  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:19

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 04/19/2016 01:45:54

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:39 |
| Air Blank          | 0.000 | 01:40 |
| Control Test       | 0.079 | 01:40 |
| Air Blank          | 0.000 | 01:40 |
| Subject Sample # 1 | 0.164 | 01:41 |
| Breath Volume      | 1.945 | 01:41 |
| Air Blank          | 0.000 | 01:41 |
| Air Blank          | 0.000 | 01:43 |
| Subject Sample # 2 | 0.161 | 01:44 |
| Breath Volume      | 2.214 | 01:44 |
| Air Blank          | 0.000 | 01:44 |
| Control Test       | 0.080 | 01:45 |
| Air Blank          | 0.000 | 01:45 |
| Diagnostics Check  | OK    | 01:45 |

---

Data Download Date/Time: 04/28/2016 10:38

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MILLER,GARY,A  
BREATH TEST OPERATOR AGENCY: Marion County SO  
ARRESTING OFFICER: MILLER,GARY  
ARRESTING OFFICER AGENCY: Marion County SO  
LAST AGENCY INSPECTION: 03/24/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 83550  
DRY GAS EXPIRY DATE: 09/26/2018  
VIOLATION CODE: DUI  
ARREST TIME: 23:52  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:44

SUBJECT'S LAST NAME: SHORES

DATE/TIME OF TEST: 04/22/2016 01:10:45

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:04 |
| Air Blank          | 0.000 | 01:04 |
| Control Test       | 0.079 | 01:05 |
| Air Blank          | 0.000 | 01:05 |
| Subject Sample # 1 | 0.116 | 01:06 |
| Breath Volume      | 2.578 | 01:06 |
| Air Blank          | 0.000 | 01:06 |
| Air Blank          | 0.000 | 01:08 |
| Subject Sample # 2 | 0.116 | 01:09 |
| Breath Volume      | 2.347 | 01:09 |
| Air Blank          | 0.000 | 01:09 |
| Control Test       | 0.082 | 01:10 |
| Air Blank          | 0.000 | 01:10 |
| Diagnostics Check  | OK    | 01:10 |

---

Data Download Date/Time: 04/28/2016 10:38

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MILLER,GARY,A  
BREATH TEST OPERATOR AGENCY: Marion County SO  
ARRESTING OFFICER: MICHAELIDES,S  
ARRESTING OFFICER AGENCY: Belleview PD  
LAST AGENCY INSPECTION: 03/24/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 83550  
DRY GAS EXPIRY DATE: 09/26/2018  
VIOLATION CODE: DUI  
ARREST TIME: 01:56  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: BLAS

DATE/TIME OF TEST: 04/23/2016 02:56:31

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:50 |
| Air Blank          | 0.000 | 02:50 |
| Control Test       | 0.080 | 02:51 |
| Air Blank          | 0.000 | 02:51 |
| Subject Sample # 1 | 0.192 | 02:52 |
| Breath Volume      | 3.601 | 02:52 |
| Air Blank          | 0.000 | 02:52 |
| Air Blank          | 0.000 | 02:54 |
| Subject Sample # 2 | 0.188 | 02:54 |
| Breath Volume      | 2.882 | 02:54 |
| Air Blank          | 0.000 | 02:55 |
| Control Test       | 0.081 | 02:55 |
| Air Blank          | 0.000 | 02:56 |
| Diagnostics Check  | OK    | 02:56 |

---

Data Download Date/Time: 04/28/2016 10:39

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DOBOSIEWICZ, ANTHONY, N  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: DOBOSIEWICZ, ANTHONY  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 03/24/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 83550  
DRY GAS EXPIRY DATE: 09/26/2018  
VIOLATION CODE: DUI  
ARREST TIME: 02:02  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: SOTO

DATE/TIME OF TEST: 04/23/2016 03:13:59

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:07 |
| Air Blank          | 0.000 | 03:07 |
| Control Test       | 0.079 | 03:08 |
| Air Blank          | 0.000 | 03:08 |
| Subject Sample # 1 | 0.164 | 03:09 |
| Breath Volume      | 1.273 | 03:09 |
| Air Blank          | 0.000 | 03:10 |
| Air Blank          | 0.000 | 03:11 |
| Subject Sample # 2 | 0.160 | 03:12 |
| Breath Volume      | 2.019 | 03:12 |
| Air Blank          | 0.000 | 03:13 |
| Control Test       | 0.080 | 03:13 |
| Air Blank          | 0.000 | 03:13 |
| Diagnostics Check  | OK    | 03:13 |

---

Data Download Date/Time: 04/28/2016 10:39

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MILLER,GARY,A  
BREATH TEST OPERATOR AGENCY: Marion County SO  
ARRESTING OFFICER: MILLER,GARY  
ARRESTING OFFICER AGENCY: Marion County SO  
LAST AGENCY INSPECTION: 03/24/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 83550  
DRY GAS EXPIRY DATE: 09/26/2018  
VIOLATION CODE: DUI  
ARREST TIME: 21:40  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:15

SUBJECT'S LAST NAME: KELLEY

DATE/TIME OF TEST: 04/23/2016 22:42:55

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:36 |
| Air Blank          | 0.000 | 22:37 |
| Control Test       | 0.080 | 22:37 |
| Air Blank          | 0.000 | 22:38 |
| Subject Sample # 1 | 0.154 | 22:38 |
| Breath Volume      | 2.972 | 22:38 |
| Air Blank          | 0.000 | 22:39 |
| Air Blank          | 0.000 | 22:40 |
| Subject Sample # 2 | 0.154 | 22:41 |
| Breath Volume      | 3.050 | 22:41 |
| Air Blank          | 0.000 | 22:42 |
| Control Test       | 0.081 | 22:42 |
| Air Blank          | 0.000 | 22:42 |
| Diagnostics Check  | OK    | 22:42 |

---

Data Download Date/Time: 04/28/2016 10:39

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DOBOSIEWICZ, ANTHONY, N  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: DOBOSIEWICZ, ANTHONY  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 03/24/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 83550  
DRY GAS EXPIRY DATE: 09/26/2018  
VIOLATION CODE: DUI  
ARREST TIME: 01:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: MARTINEZ LEON

DATE/TIME OF TEST: 04/24/2016 02:16:32

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:10 |
| Air Blank          | 0.000 | 02:10 |
| Control Test       | 0.079 | 02:11 |
| Air Blank          | 0.000 | 02:11 |
| Subject Sample # 1 | 0.099 | 02:12 |
| Breath Volume      | 3.949 | 02:12 |
| Air Blank          | 0.000 | 02:12 |
| Air Blank          | 0.000 | 02:14 |
| Subject Sample # 2 | 0.090 | 02:15 |
| Breath Volume      | 2.921 | 02:15 |
| Air Blank          | 0.000 | 02:15 |
| Control Test       | 0.081 | 02:15 |
| Air Blank          | 0.000 | 02:16 |
| Diagnostics Check  | OK    | 02:16 |

---

Data Download Date/Time: 04/28/2016 10:39



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LETSON,REBECCA,G  
BREATH TEST OPERATOR AGENCY: Ocala PD  
ARRESTING OFFICER: LETSON,REBECCA  
ARRESTING OFFICER AGENCY: Ocala PD  
LAST AGENCY INSPECTION: 03/24/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 83550  
DRY GAS EXPIRY DATE: 09/26/2018  
VIOLATION CODE: DUI  
ARREST TIME: 02:51  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: PAVICIC

DATE/TIME OF TEST: 04/24/2016 03:34:01

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:27 |
| Air Blank          | 0.000 | 03:27 |
| Control Test       | 0.080 | 03:28 |
| Air Blank          | 0.000 | 03:28 |
| Subject Sample # 1 | 0.195 | 03:29 |
| Breath Volume      | 2.039 | 03:29 |
| Air Blank          | 0.000 | 03:30 |
| Air Blank          | 0.000 | 03:31 |
| Subject Sample # 2 | 0.196 | 03:32 |
| Breath Volume      | 1.925 | 03:32 |
| Air Blank          | 0.000 | 03:33 |
| Control Test       | 0.081 | 03:33 |
| Air Blank          | 0.000 | 03:33 |
| Diagnostics Check  | OK    | 03:34 |

---

Data Download Date/Time: 04/28/2016 10:39

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GAGO-  
CHUQUILLANQUI,ALBERTO,E  
BREATH TEST OPERATOR AGENCY: Marion County SO  
ARRESTING OFFICER: GAGO-  
CHUQUILLANQUI,ALBERTO  
ARRESTING OFFICER AGENCY: Marion County SO  
LAST AGENCY INSPECTION: 03/24/2016

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 83550  
DRY GAS EXPIRY DATE: 09/26/2018  
  
VIOLATION CODE: DUI  
ARREST TIME: 23:38  
GENDER: Male  
  
OBSERVATION PERIOD BEGAN: 00:06

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 04/26/2016 00:33:04

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:26 |
| Air Blank          | 0.000 | 00:27 |
| Control Test       | 0.080 | 00:27 |
| Air Blank          | 0.000 | 00:27 |
| Subject Sample # 1 | 0.116 | 00:28 |
| Breath Volume      | 1.738 | 00:28 |
| Air Blank          | 0.000 | 00:29 |
| Air Blank          | 0.000 | 00:31 |
| Subject Sample # 2 | 0.109 | 00:31 |
| Breath Volume      | 1.148 | 00:31 |
| Air Blank          | 0.000 | 00:32 |
| Control Test       | 0.079 | 00:32 |
| Air Blank          | 0.000 | 00:32 |
| Diagnostics Check  | OK    | 00:33 |

---

Data Download Date/Time: 04/28/2016 10:39

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WRIGHT,JUSTIN,P  
BREATH TEST OPERATOR AGENCY: Sumter County SO  
ARRESTING OFFICER: MATA,ROBERTH  
ARRESTING OFFICER AGENCY: Sumter County SO  
LAST AGENCY INSPECTION: 03/28/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28814080A3  
DRY GAS EXPIRY DATE: 11/05/2016  
VIOLATION CODE: DUI  
ARREST TIME: 21:31  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:08

SUBJECT'S LAST NAME: LANDRY

DATE/TIME OF TEST: 04/03/2016 22:38:02

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 22:35 |                      |
| Air Blank         | 0.000 | 22:35 |                      |
| Control Test      | 0.077 | 22:35 |                      |
| Air Blank         | 0.000 | 22:36 |                      |
| Subject Sample    | # 1   | 22:36 | Subject Test Refused |
| Breath Volume     | 0.000 | 22:36 |                      |
| Air Blank         | 0.000 | 22:37 |                      |
| Control Test      | 0.077 | 22:37 |                      |
| Air Blank         | 0.000 | 22:37 |                      |
| Diagnostics Check | OK    | 22:38 |                      |

---

Data Download Date/Time: 04/24/2016 18:36

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JOHNSON,BRENDA,R  
BREATH TEST OPERATOR AGENCY: Sumter County SO  
ARRESTING OFFICER: REYNOLDS,MATTHEWS  
ARRESTING OFFICER AGENCY: Wildwood PD  
LAST AGENCY INSPECTION: 03/28/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28814080A3  
DRY GAS EXPIRY DATE: 11/05/2016  
VIOLATION CODE: DUI  
ARREST TIME: 21:34  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: HUTTEL

DATE/TIME OF TEST: 04/08/2016 22:44:33

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:37 |
| Air Blank          | 0.000 | 22:37 |
| Control Test       | 0.077 | 22:38 |
| Air Blank          | 0.000 | 22:38 |
| Subject Sample # 1 | 0.127 | 22:39 |
| Breath Volume      | 2.128 | 22:39 |
| Air Blank          | 0.000 | 22:40 |
| Air Blank          | 0.000 | 22:42 |
| Subject Sample # 2 | 0.129 | 22:42 |
| Breath Volume      | 1.992 | 22:42 |
| Air Blank          | 0.000 | 22:43 |
| Control Test       | 0.077 | 22:43 |
| Air Blank          | 0.000 | 22:44 |
| Diagnostics Check  | OK    | 22:44 |

---

Data Download Date/Time: 04/24/2016 18:36

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WIGGLESWORTH,JOE,B  
BREATH TEST OPERATOR AGENCY: Sumter County SO  
ARRESTING OFFICER: GALBREATH,D  
ARRESTING OFFICER AGENCY: Sumter County SO  
LAST AGENCY INSPECTION: 03/28/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28814080A3  
DRY GAS EXPIRY DATE: 11/05/2016  
VIOLATION CODE: DUI  
ARREST TIME: 22:59  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: SIMMONS

DATE/TIME OF TEST: 04/10/2016 23:59:06

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:52 |
| Air Blank          | 0.000 | 23:52 |
| Control Test       | 0.077 | 23:53 |
| Air Blank          | 0.000 | 23:53 |
| Subject Sample # 1 | 0.152 | 23:54 |
| Breath Volume      | 1.816 | 23:54 |
| Air Blank          | 0.000 | 23:54 |
| Air Blank          | 0.000 | 23:56 |
| Subject Sample # 2 | 0.144 | 23:57 |
| Breath Volume      | 2.062 | 23:57 |
| Air Blank          | 0.000 | 23:58 |
| Control Test       | 0.077 | 23:58 |
| Air Blank          | 0.000 | 23:59 |
| Diagnostics Check  | OK    | 23:59 |

---

Data Download Date/Time: 04/24/2016 18:36

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SMITH,TRAVIS,D  
BREATH TEST OPERATOR AGENCY: Sumter County SO  
ARRESTING OFFICER: TAN,RICHARD  
ARRESTING OFFICER AGENCY: Sumter County SO  
LAST AGENCY INSPECTION: 03/28/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28814080A3  
DRY GAS EXPIRY DATE: 11/05/2016  
VIOLATION CODE: DUI  
ARREST TIME: 02:16  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:46

SUBJECT'S LAST NAME: BASFORD

DATE/TIME OF TEST: 04/12/2016 03:10:59

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 03:07 |                      |
| Air Blank         | 0.000 | 03:08 |                      |
| Control Test      | 0.077 | 03:08 |                      |
| Air Blank         | 0.000 | 03:09 |                      |
| Subject Sample    | # 1   | 03:09 | Subject Test Refused |
| Breath Volume     | 0.000 | 03:09 |                      |
| Air Blank         | 0.000 | 03:10 |                      |
| Control Test      | 0.076 | 03:10 |                      |
| Air Blank         | 0.000 | 03:10 |                      |
| Diagnostics Check | OK    | 03:10 |                      |

---

Data Download Date/Time: 04/24/2016 18:36

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LANCE,RAY,C  
BREATH TEST OPERATOR AGENCY: Sumter County SO  
ARRESTING OFFICER: OTERIO,HECTOR  
ARRESTING OFFICER AGENCY: Sumter County SO  
LAST AGENCY INSPECTION: 03/28/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28814080A3  
DRY GAS EXPIRY DATE: 11/05/2016  
VIOLATION CODE: DUI  
ARREST TIME: 23:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: FUSCHINI

DATE/TIME OF TEST: 04/15/2016 23:55:07

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 23:52 |                      |
| Air Blank         | 0.000 | 23:52 |                      |
| Control Test      | 0.076 | 23:52 |                      |
| Air Blank         | 0.000 | 23:53 |                      |
| Subject Sample    | # 1   | 23:53 | Subject Test Refused |
| Breath Volume     | 0.000 | 23:53 |                      |
| Air Blank         | 0.000 | 23:54 |                      |
| Control Test      | 0.077 | 23:54 |                      |
| Air Blank         | 0.000 | 23:55 |                      |
| Diagnostics Check | OK    | 23:55 |                      |

---

Data Download Date/Time: 04/24/2016 18:36

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HOLLIS,TERENCE,L  
BREATH TEST OPERATOR AGENCY: Sumter County SO  
ARRESTING OFFICER: GARTNER,A  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 03/28/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28814080A3  
DRY GAS EXPIRY DATE: 11/05/2016  
VIOLATION CODE: DUI  
ARREST TIME: 23:32  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:08

SUBJECT'S LAST NAME: ANDREWS

DATE/TIME OF TEST: 04/20/2016 00:38:02

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:28 |
| Air Blank          | 0.000 | 00:29 |
| Control Test       | 0.077 | 00:29 |
| Air Blank          | 0.000 | 00:30 |
| Subject Sample # 1 | 0.128 | 00:31 |
| Breath Volume      | 1.453 | 00:31 |
| Air Blank          | 0.000 | 00:31 |
| Air Blank          | 0.000 | 00:33 |
| Subject Sample # 2 | 0.127 | 00:36 |
| Breath Volume      | 1.253 | 00:36 |
| Air Blank          | 0.000 | 00:37 |
| Control Test       | 0.079 | 00:37 |
| Air Blank          | 0.000 | 00:37 |
| Diagnostics Check  | OK    | 00:38 |

---

Data Download Date/Time: 04/24/2016 18:36



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LANCE,RAY,C  
BREATH TEST OPERATOR AGENCY: Sumter County SO  
ARRESTING OFFICER: MCCAIG,ASHLEY  
ARRESTING OFFICER AGENCY: Sumter County SO  
LAST AGENCY INSPECTION: 03/28/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28814080A3  
DRY GAS EXPIRY DATE: 11/05/2016  
VIOLATION CODE: DUI  
ARREST TIME: 20:09  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:40

SUBJECT'S LAST NAME: JENNINGS

DATE/TIME OF TEST: 04/21/2016 21:07:33

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 21:00 |
| Air Blank          | 0.000 | 21:01 |
| Control Test       | 0.077 | 21:01 |
| Air Blank          | 0.000 | 21:02 |
| Subject Sample # 1 | 0.112 | 21:02 |
| Breath Volume      | 3.289 | 21:02 |
| Air Blank          | 0.000 | 21:03 |
| Air Blank          | 0.000 | 21:05 |
| Subject Sample # 2 | 0.110 | 21:05 |
| Breath Volume      | 3.027 | 21:05 |
| Air Blank          | 0.000 | 21:06 |
| Control Test       | 0.076 | 21:06 |
| Air Blank          | 0.000 | 21:07 |
| Diagnostics Check  | OK    | 21:07 |

---

Data Download Date/Time: 04/24/2016 18:36

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SIBILIA, CHRISTOPHER, R  
BREATH TEST OPERATOR AGENCY: CCA Citrus  
ARRESTING OFFICER: VICK, B  
ARRESTING OFFICER AGENCY: Citrus County SO  
LAST AGENCY INSPECTION: 03/18/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13115080A1  
DRY GAS EXPIRY DATE: 06/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 18:23  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:06

SUBJECT'S LAST NAME: BLANCHE

DATE/TIME OF TEST: 04/02/2016 19:51:18

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 19:43 |
| Air Blank          | 0.000 | 19:44 |
| Control Test       | 0.077 | 19:44 |
| Air Blank          | 0.000 | 19:44 |
| Subject Sample # 1 | 0.176 | 19:46 |
| Breath Volume      | 1.617 | 19:46 |
| Air Blank          | 0.000 | 19:47 |
| Air Blank          | 0.000 | 19:48 |
| Subject Sample # 2 | 0.173 | 19:49 |
| Breath Volume      | 1.222 | 19:49 |
| Air Blank          | 0.000 | 19:50 |
| Control Test       | 0.078 | 19:50 |
| Air Blank          | 0.000 | 19:51 |
| Diagnostics Check  | OK    | 19:51 |

---

Data Download Date/Time: 04/28/2016 11:51

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RHODE,ERIC,C  
BREATH TEST OPERATOR AGENCY: CCA Citrus  
ARRESTING OFFICER: REID,W  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 03/18/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13115080A1  
DRY GAS EXPIRY DATE: 06/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 23:32  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: MARTINDALE

DATE/TIME OF TEST: 04/03/2016 00:58:04

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:51 |
| Air Blank          | 0.000 | 00:52 |
| Control Test       | 0.077 | 00:52 |
| Air Blank          | 0.000 | 00:52 |
| Subject Sample # 1 | 0.067 | 00:53 |
| Breath Volume      | 2.113 | 00:53 |
| Air Blank          | 0.000 | 00:54 |
| Air Blank          | 0.000 | 00:55 |
| Subject Sample # 2 | 0.074 | 00:56 |
| Breath Volume      | 1.539 | 00:56 |
| Air Blank          | 0.000 | 00:57 |
| Control Test       | 0.078 | 00:57 |
| Air Blank          | 0.000 | 00:57 |
| Diagnostics Check  | OK    | 00:58 |

---

Data Download Date/Time: 04/28/2016 11:51

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RHODE,ERIC,C  
BREATH TEST OPERATOR AGENCY: CCA Citrus  
ARRESTING OFFICER: TITUS,J  
ARRESTING OFFICER AGENCY: Citrus County SO  
LAST AGENCY INSPECTION: 03/18/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13115080A1  
DRY GAS EXPIRY DATE: 06/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 18:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:00

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 04/03/2016 19:36:53

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 19:26 |
| Air Blank          | 0.000 | 19:27 |
| Control Test       | 0.077 | 19:27 |
| Air Blank          | 0.000 | 19:27 |
| Subject Sample # 1 | 0.305 | 19:29 |
| Breath Volume      | 1.937 | 19:29 |
| Air Blank          | 0.000 | 19:29 |
| Air Blank          | 0.000 | 19:31 |
| Subject Sample # 2 | 0.344 | 19:32 |
| Breath Volume      | 2.878 | 19:32 |
| Air Blank          | 0.000 | 19:32 |
| Air Blank          | 0.000 | 19:34 |
| Subject Sample # 3 | 0.328 | 19:35 |
| Breath Volume      | 2.796 | 19:35 |
| Air Blank          | 0.000 | 19:35 |
| Control Test       | 0.076 | 19:36 |
| Air Blank          | 0.000 | 19:36 |
| Diagnostics Check  | OK    | 19:36 |

---

Data Download Date/Time: 04/28/2016 11:51

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BOGUE,VINCE,H  
BREATH TEST OPERATOR AGENCY: CCA Citrus  
ARRESTING OFFICER: BLOTZ,G  
ARRESTING OFFICER AGENCY: Citrus County SO  
LAST AGENCY INSPECTION: 03/18/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13115080A1  
DRY GAS EXPIRY DATE: 06/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 22:39  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: DAY

DATE/TIME OF TEST: 04/07/2016 23:46:30

---

|                   |           |       |                 |
|-------------------|-----------|-------|-----------------|
| Diagnostics Check | OK        | 23:34 |                 |
| Air Blank         | 0.000     | 23:35 |                 |
| Control Test      | 0.077     | 23:35 |                 |
| Air Blank         | 0.000     | 23:36 |                 |
| Subject Sample    | # 1 0.207 | 23:39 | Volume Not Met  |
| Breath Volume     | 0.832     | 23:39 |                 |
| Air Blank         | 0.000     | 23:40 |                 |
| Air Blank         | 0.000     | 23:41 |                 |
| Subject Sample    | # 2 0.206 | 23:44 | Slope Not Level |
| Breath Volume     | 1.156     | 23:44 |                 |
| Air Blank         | 0.000     | 23:45 |                 |
| Control Test      | 0.075     | 23:45 |                 |
| Air Blank         | 0.000     | 23:46 |                 |
| Diagnostics Check | OK        | 23:46 |                 |

---

Data Download Date/Time: 04/28/2016 11:51

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BOGUE,VINCE,H  
BREATH TEST OPERATOR AGENCY: CCA Citrus  
ARRESTING OFFICER: BLOTZ,G  
ARRESTING OFFICER AGENCY: Citrus County SO  
LAST AGENCY INSPECTION: 03/18/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13115080A1  
DRY GAS EXPIRY DATE: 06/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 22:39  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: DAY

DATE/TIME OF TEST: 04/08/2016 00:31:09

---

|                   |           |       |                      |
|-------------------|-----------|-------|----------------------|
| Diagnostics Check | OK        | 00:17 |                      |
| Air Blank         | 0.000     | 00:17 |                      |
| Control Test      | 0.076     | 00:17 |                      |
| Air Blank         | 0.000     | 00:18 |                      |
| Subject Sample    | # 1 0.201 | 00:19 |                      |
| Breath Volume     | 1.148     | 00:19 |                      |
| Air Blank         | 0.000     | 00:20 |                      |
| Air Blank         | 0.000     | 00:22 |                      |
| Subject Sample    | # 2 0.199 | 00:25 | Volume Not Met       |
| Breath Volume     | 0.371     | 00:25 |                      |
| Air Blank         | 0.000     | 00:26 |                      |
| Air Blank         | 0.000     | 00:27 |                      |
| Subject Sample    | # 3       | 00:29 | Subject Test Refused |
| Breath Volume     | 0.910     | 00:29 |                      |
| Air Blank         | 0.000     | 00:30 |                      |
| Control Test      | 0.076     | 00:30 |                      |
| Air Blank         | 0.000     | 00:31 |                      |
| Diagnostics Check | OK        | 00:31 |                      |

---

Data Download Date/Time: 04/28/2016 11:51

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MCCAZZIO,SEAN,J  
BREATH TEST OPERATOR AGENCY: CCA Citrus  
ARRESTING OFFICER: BARR,J  
ARRESTING OFFICER AGENCY: Citrus County SO  
LAST AGENCY INSPECTION: 03/18/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13115080A1  
DRY GAS EXPIRY DATE: 06/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 01:23  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: N632110563830

DATE/TIME OF TEST: 04/10/2016 02:11:50

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 02:09 |                      |
| Air Blank         | 0.000 | 02:09 |                      |
| Control Test      | 0.078 | 02:09 |                      |
| Air Blank         | 0.000 | 02:10 |                      |
| Subject Sample    | # 1   | 02:10 | Subject Test Refused |
| Breath Volume     | 0.000 | 02:10 |                      |
| Air Blank         | 0.000 | 02:10 |                      |
| Control Test      | 0.078 | 02:11 |                      |
| Air Blank         | 0.000 | 02:11 |                      |
| Diagnostics Check | OK    | 02:11 |                      |

---

Data Download Date/Time: 04/28/2016 11:51

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: METZ,KYLE,J  
BREATH TEST OPERATOR AGENCY: CCA Citrus  
ARRESTING OFFICER: COUNTY,CITRUS  
ARRESTING OFFICER AGENCY: CCA Citrus  
LAST AGENCY INSPECTION: 03/18/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13115080A1  
DRY GAS EXPIRY DATE: 06/05/2017  
VIOLATION CODE: ZERO TOL  
ARREST TIME: 08:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 08:05

SUBJECT'S LAST NAME: BAGLEY

DATE/TIME OF TEST: 04/14/2016 08:31:27

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 08:25 |
| Air Blank          | 0.000 | 08:25 |
| Control Test       | 0.078 | 08:26 |
| Air Blank          | 0.000 | 08:26 |
| Subject Sample # 1 | 0.000 | 08:27 |
| Breath Volume      | 3.714 | 08:27 |
| Air Blank          | 0.000 | 08:27 |
| Air Blank          | 0.000 | 08:29 |
| Subject Sample # 2 | 0.000 | 08:30 |
| Breath Volume      | 3.230 | 08:30 |
| Air Blank          | 0.000 | 08:30 |
| Control Test       | 0.076 | 08:30 |
| Air Blank          | 0.000 | 08:31 |
| Diagnostics Check  | OK    | 08:31 |

---

Data Download Date/Time: 04/28/2016 11:51



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: METZ,KYLE,J  
BREATH TEST OPERATOR AGENCY: CCA Citrus  
ARRESTING OFFICER: COUNTY,CITRUS  
ARRESTING OFFICER AGENCY: CCA Citrus  
LAST AGENCY INSPECTION: 03/18/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13115080A1  
DRY GAS EXPIRY DATE: 06/05/2017  
VIOLATION CODE: ZERO TOL  
ARREST TIME: 08:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 08:30

SUBJECT'S LAST NAME: ALLAIRE

DATE/TIME OF TEST: 04/14/2016 08:56:08

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 08:50 |
| Air Blank          | 0.000 | 08:50 |
| Control Test       | 0.077 | 08:51 |
| Air Blank          | 0.000 | 08:51 |
| Subject Sample # 1 | 0.000 | 08:52 |
| Breath Volume      | 3.144 | 08:52 |
| Air Blank          | 0.000 | 08:52 |
| Air Blank          | 0.000 | 08:54 |
| Subject Sample # 2 | 0.000 | 08:54 |
| Breath Volume      | 3.179 | 08:54 |
| Air Blank          | 0.000 | 08:55 |
| Control Test       | 0.076 | 08:55 |
| Air Blank          | 0.000 | 08:56 |
| Diagnostics Check  | OK    | 08:56 |

---

Data Download Date/Time: 04/28/2016 11:51

# **Florida Department of Law Enforcement**

## **Alcohol Testing Program**

### **Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: METZ,KYLE,W  
BREATH TEST OPERATOR AGENCY: CCA Citrus  
ARRESTING OFFICER: COUNTY,CITRUS  
ARRESTING OFFICER AGENCY: CCA Citrus  
LAST AGENCY INSPECTION: 03/18/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13115080A1  
DRY GAS EXPIRY DATE: 06/05/2017  
VIOLATION CODE: ZERO TOL  
ARREST TIME: 08:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 08:10

SUBJECT'S LAST NAME: BARGIEL

DATE/TIME OF TEST: 04/14/2016 09:18:02

---

|                   |       |       |                  |
|-------------------|-------|-------|------------------|
| Diagnostics Check | OK    | 09:17 |                  |
| Air Blank         |       | 09:17 | Sequence Aborted |
| Air Blank         | 0.000 | 09:18 |                  |

---

Data Download Date/Time: 04/28/2016 11:51

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: METZ,KYLE,J  
BREATH TEST OPERATOR AGENCY: CCA Citrus  
ARRESTING OFFICER: COUNTY,CITRUS  
ARRESTING OFFICER AGENCY: CCA Citrus  
LAST AGENCY INSPECTION: 03/18/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13115080A1  
DRY GAS EXPIRY DATE: 06/05/2017  
VIOLATION CODE: ZERO TOL  
ARREST TIME: 09:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 09:10

SUBJECT'S LAST NAME: BARGIEL

DATE/TIME OF TEST: 04/14/2016 09:36:25

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 09:30 |
| Air Blank          | 0.000 | 09:30 |
| Control Test       | 0.078 | 09:31 |
| Air Blank          | 0.000 | 09:31 |
| Subject Sample # 1 | 0.000 | 09:32 |
| Breath Volume      | 2.093 | 09:32 |
| Air Blank          | 0.000 | 09:32 |
| Air Blank          | 0.000 | 09:34 |
| Subject Sample # 2 | 0.000 | 09:35 |
| Breath Volume      | 1.308 | 09:35 |
| Air Blank          | 0.000 | 09:35 |
| Control Test       | 0.077 | 09:35 |
| Air Blank          | 0.000 | 09:36 |
| Diagnostics Check  | OK    | 09:36 |

---

Data Download Date/Time: 04/28/2016 11:51

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WING,JOSHUA,M  
BREATH TEST OPERATOR AGENCY: CCA Citrus  
ARRESTING OFFICER: ZAREK,MICHAEL  
ARRESTING OFFICER AGENCY: Citrus County SO  
LAST AGENCY INSPECTION: 03/18/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13115080A1  
DRY GAS EXPIRY DATE: 06/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 14:08  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 15:16

SUBJECT'S LAST NAME: MAY

DATE/TIME OF TEST: 04/15/2016 15:40:26

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 15:37 |                      |
| Air Blank         | 0.000 | 15:37 |                      |
| Control Test      | 0.077 | 15:38 |                      |
| Air Blank         | 0.000 | 15:38 |                      |
| Subject Sample    | # 1   | 15:39 | Subject Test Refused |
| Breath Volume     | 0.000 | 15:39 |                      |
| Air Blank         | 0.000 | 15:39 |                      |
| Control Test      | 0.077 | 15:39 |                      |
| Air Blank         | 0.000 | 15:40 |                      |
| Diagnostics Check | OK    | 15:40 |                      |

---

Data Download Date/Time: 04/28/2016 11:51

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SIBILIA,CHRISTOPHER,R  
BREATH TEST OPERATOR AGENCY: CCA Citrus  
ARRESTING OFFICER: BLOTZ,J  
ARRESTING OFFICER AGENCY: Citrus County SO  
LAST AGENCY INSPECTION: 03/18/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13115080A1  
DRY GAS EXPIRY DATE: 06/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 19:21  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 20:13

SUBJECT'S LAST NAME: LAND

DATE/TIME OF TEST: 04/15/2016 20:54:36

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 20:46 |
| Air Blank          | 0.000 | 20:47 |
| Control Test       | 0.077 | 20:47 |
| Air Blank          | 0.000 | 20:48 |
| Subject Sample # 1 | 0.158 | 20:49 |
| Breath Volume      | 1.367 | 20:49 |
| Air Blank          | 0.000 | 20:50 |
| Air Blank          | 0.000 | 20:52 |
| Subject Sample # 2 | 0.156 | 20:53 |
| Breath Volume      | 1.496 | 20:53 |
| Air Blank          | 0.000 | 20:53 |
| Control Test       | 0.077 | 20:54 |
| Air Blank          | 0.000 | 20:54 |
| Diagnostics Check  | OK    | 20:54 |

---

Data Download Date/Time: 04/28/2016 11:51

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RHODE,ERIC,C  
BREATH TEST OPERATOR AGENCY: CCA Citrus  
ARRESTING OFFICER: BLOTZ,G  
ARRESTING OFFICER AGENCY: Citrus County SO  
LAST AGENCY INSPECTION: 03/18/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13115080A1  
DRY GAS EXPIRY DATE: 06/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 22:44  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:27

SUBJECT'S LAST NAME: MICHIE

DATE/TIME OF TEST: 04/17/2016 00:11:51

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:05 |
| Air Blank          | 0.000 | 00:05 |
| Control Test       | 0.078 | 00:06 |
| Air Blank          | 0.000 | 00:06 |
| Subject Sample # 1 | 0.000 | 00:07 |
| Breath Volume      | 2.273 | 00:07 |
| Air Blank          | 0.000 | 00:07 |
| Air Blank          | 0.000 | 00:09 |
| Subject Sample # 2 | 0.000 | 00:10 |
| Breath Volume      | 2.667 | 00:10 |
| Air Blank          | 0.000 | 00:10 |
| Control Test       | 0.078 | 00:11 |
| Air Blank          | 0.000 | 00:11 |
| Diagnostics Check  | OK    | 00:11 |

---

Data Download Date/Time: 04/28/2016 11:51

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SCHOETTNER,ADRIAN,J  
BREATH TEST OPERATOR AGENCY: CCA Citrus  
ARRESTING OFFICER: BLOTZ,G  
ARRESTING OFFICER AGENCY: Citrus County SO  
LAST AGENCY INSPECTION: 03/18/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13115080A1  
DRY GAS EXPIRY DATE: 06/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 23:48  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: BILLINGS

DATE/TIME OF TEST: 04/18/2016 00:33:47

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 00:30 |                      |
| Air Blank         | 0.000 | 00:31 |                      |
| Control Test      | 0.077 | 00:31 |                      |
| Air Blank         | 0.000 | 00:32 |                      |
| Subject Sample    | # 1   | 00:32 | Subject Test Refused |
| Breath Volume     | 0.000 | 00:32 |                      |
| Air Blank         | 0.000 | 00:32 |                      |
| Control Test      | 0.077 | 00:33 |                      |
| Air Blank         | 0.000 | 00:33 |                      |
| Diagnostics Check | OK    | 00:33 |                      |

---

Data Download Date/Time: 04/28/2016 11:51

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: METZ,KYLE,J  
BREATH TEST OPERATOR AGENCY: CCA Citrus  
ARRESTING OFFICER: SUDLOW,TROOPER  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 03/18/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13115080A1  
DRY GAS EXPIRY DATE: 06/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 16:09  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:44

SUBJECT'S LAST NAME: MANCINI

DATE/TIME OF TEST: 04/18/2016 17:11:23

---

|                   |           |       |                           |
|-------------------|-----------|-------|---------------------------|
| Diagnostics Check | OK        | 17:04 |                           |
| Air Blank         | 0.000     | 17:05 |                           |
| Control Test      | 0.077     | 17:05 |                           |
| Air Blank         | 0.000     | 17:06 |                           |
| Subject Sample    | # 1 0.195 | 17:06 |                           |
| Breath Volume     | 3.511     | 17:06 |                           |
| Air Blank         | 0.000     | 17:07 |                           |
| Air Blank         | 0.000     | 17:09 |                           |
| Subject Sample    | # 2 0.180 | 17:09 |                           |
| Breath Volume     | 1.660     | 17:09 |                           |
| Air Blank         | 0.000     | 17:10 |                           |
| Control Test      | 0.073     | 17:10 | Control Outside Tolerance |
| Air Blank         | 0.000     | 17:11 |                           |

---

Data Download Date/Time: 04/28/2016 11:51



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RHODE,ERIC,C  
BREATH TEST OPERATOR AGENCY: CCA Citrus  
ARRESTING OFFICER: TITUS,J  
ARRESTING OFFICER AGENCY: Citrus County SO  
LAST AGENCY INSPECTION: 03/18/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13115080A1  
DRY GAS EXPIRY DATE: 06/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 03:44  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:18

SUBJECT'S LAST NAME: DYKES

DATE/TIME OF TEST: 04/22/2016 04:54:01

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:46 |
| Air Blank          | 0.000 | 04:46 |
| Control Test       | 0.077 | 04:46 |
| Air Blank          | 0.000 | 04:47 |
| Subject Sample # 1 | 0.186 | 04:48 |
| Breath Volume      | 3.027 | 04:48 |
| Air Blank          | 0.000 | 04:49 |
| Air Blank          | 0.000 | 04:51 |
| Subject Sample # 2 | 0.190 | 04:52 |
| Breath Volume      | 3.621 | 04:52 |
| Air Blank          | 0.000 | 04:53 |
| Control Test       | 0.076 | 04:53 |
| Air Blank          | 0.000 | 04:53 |
| Diagnostics Check  | OK    | 04:54 |

---

Data Download Date/Time: 04/28/2016 11:51

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PARKER,PETER,  
BREATH TEST OPERATOR AGENCY: CCA Citrus  
ARRESTING OFFICER: GERENA,G  
ARRESTING OFFICER AGENCY: Citrus County SO  
LAST AGENCY INSPECTION: 03/18/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13115080A1  
DRY GAS EXPIRY DATE: 06/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 03:04  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:26

SUBJECT'S LAST NAME: SCHNEIDER

DATE/TIME OF TEST: 04/23/2016 04:18:39

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:12 |
| Air Blank          | 0.000 | 04:12 |
| Control Test       | 0.077 | 04:12 |
| Air Blank          | 0.000 | 04:13 |
| Subject Sample # 1 | 0.218 | 04:13 |
| Breath Volume      | 2.281 | 04:13 |
| Air Blank          | 0.000 | 04:14 |
| Air Blank          | 0.000 | 04:16 |
| Subject Sample # 2 | 0.217 | 04:17 |
| Breath Volume      | 2.843 | 04:17 |
| Air Blank          | 0.000 | 04:17 |
| Control Test       | 0.075 | 04:18 |
| Air Blank          | 0.000 | 04:18 |
| Diagnostics Check  | OK    | 04:18 |

---

Data Download Date/Time: 04/28/2016 11:51

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WING,JOSHUA,M  
BREATH TEST OPERATOR AGENCY: CCA Citrus  
ARRESTING OFFICER: COLEMAN,B  
ARRESTING OFFICER AGENCY: Citrus County SO  
LAST AGENCY INSPECTION: 03/18/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13115080A1  
DRY GAS EXPIRY DATE: 06/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 09:48  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 10:34

SUBJECT'S LAST NAME: AULTMAN

DATE/TIME OF TEST: 04/26/2016 11:01:11

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 10:54 |
| Air Blank          | 0.000 | 10:55 |
| Control Test       | 0.076 | 10:55 |
| Air Blank          | 0.000 | 10:56 |
| Subject Sample # 1 | 0.248 | 10:56 |
| Breath Volume      | 2.000 | 10:56 |
| Air Blank          | 0.000 | 10:57 |
| Air Blank          | 0.000 | 10:59 |
| Subject Sample # 2 | 0.252 | 10:59 |
| Breath Volume      | 2.117 | 10:59 |
| Air Blank          | 0.000 | 11:00 |
| Control Test       | 0.077 | 11:00 |
| Air Blank          | 0.000 | 11:01 |
| Diagnostics Check  | OK    | 11:01 |

---

Data Download Date/Time: 04/28/2016 11:51

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RHODE,ERIC,C  
BREATH TEST OPERATOR AGENCY: CCA Citrus  
ARRESTING OFFICER: TITUS,J  
ARRESTING OFFICER AGENCY: Citrus County SO  
LAST AGENCY INSPECTION: 03/18/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13115080A1  
DRY GAS EXPIRY DATE: 06/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 22:49  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:58

SUBJECT'S LAST NAME: VERDEJA LARA

DATE/TIME OF TEST: 04/27/2016 00:39:07

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 00:36 |                      |
| Air Blank         | 0.000 | 00:36 |                      |
| Control Test      | 0.076 | 00:37 |                      |
| Air Blank         | 0.000 | 00:37 |                      |
| Subject Sample    | # 1   | 00:37 | Subject Test Refused |
| Breath Volume     | 0.000 | 00:37 |                      |
| Air Blank         | 0.000 | 00:38 |                      |
| Control Test      | 0.077 | 00:38 |                      |
| Air Blank         | 0.000 | 00:39 |                      |
| Diagnostics Check | OK    | 00:39 |                      |

---

Data Download Date/Time: 04/28/2016 11:51

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MCCAZZIO,SEAN,J  
BREATH TEST OPERATOR AGENCY: CCA Citrus  
ARRESTING OFFICER: BRIGGS,R  
ARRESTING OFFICER AGENCY: Citrus County SO  
LAST AGENCY INSPECTION: 03/18/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13115080A1  
DRY GAS EXPIRY DATE: 06/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 19:53  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 20:10

SUBJECT'S LAST NAME: BEAM

DATE/TIME OF TEST: 04/27/2016 20:34:56

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 20:32 |                      |
| Air Blank         | 0.000 | 20:32 |                      |
| Control Test      | 0.077 | 20:32 |                      |
| Air Blank         | 0.000 | 20:33 |                      |
| Subject Sample    | # 1   | 20:33 | Subject Test Refused |
| Breath Volume     | 0.000 | 20:33 |                      |
| Air Blank         | 0.000 | 20:34 |                      |
| Control Test      | 0.076 | 20:34 |                      |
| Air Blank         | 0.000 | 20:34 |                      |
| Diagnostics Check | OK    | 20:34 |                      |

---

Data Download Date/Time: 04/28/2016 11:51

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: AKINS,GARY,G  
BREATH TEST OPERATOR AGENCY: FHP Troop D  
ARRESTING OFFICER: AKINS,GARY  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 03/31/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1766638  
DRY GAS EXPIRY DATE: 11/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 20:43  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: CRUZ

DATE/TIME OF TEST: 04/10/2016 21:46:58

---

|                   |       |       |            |
|-------------------|-------|-------|------------|
| Diagnostics Check | OK    | 21:45 |            |
| Air Blank         | 0.000 | 21:46 |            |
| Control Test      | 0.000 | 21:46 | RFI Detect |
| Air Blank         | 0.000 | 21:46 |            |

---

Data Download Date/Time: 04/30/2016 20:11

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: AKINS,GARY,G  
BREATH TEST OPERATOR AGENCY: FHP Troop D  
ARRESTING OFFICER: AKINS,GARY  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 03/31/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1766638  
DRY GAS EXPIRY DATE: 11/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 20:43  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: CRUZ

DATE/TIME OF TEST: 04/10/2016 21:58:43

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 21:49 |
| Air Blank          | 0.000 | 21:49 |
| Control Test       | 0.078 | 21:50 |
| Air Blank          | 0.000 | 21:50 |
| Subject Sample # 1 | 0.192 | 21:51 |
| Breath Volume      | 2.062 | 21:51 |
| Air Blank          | 0.000 | 21:51 |
| Air Blank          | 0.000 | 21:53 |
| Subject Sample # 2 | 0.171 | 21:54 |
| Breath Volume      | 1.230 | 21:54 |
| Air Blank          | 0.000 | 21:54 |
| Air Blank          | 0.000 | 21:56 |
| Subject Sample # 3 | 0.176 | 21:57 |
| Breath Volume      | 3.410 | 21:57 |
| Air Blank          | 0.000 | 21:57 |
| Control Test       | 0.078 | 21:58 |
| Air Blank          | 0.000 | 21:58 |
| Diagnostics Check  | OK    | 21:58 |

---

Data Download Date/Time: 04/30/2016 20:11

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: AKINS,GARY,G  
BREATH TEST OPERATOR AGENCY: FHP Troop D  
ARRESTING OFFICER: AKINS,GARY  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 03/31/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1766638  
DRY GAS EXPIRY DATE: 11/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 19:46  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 19:50

SUBJECT'S LAST NAME: SPEARMAN-HOWARD

DATE/TIME OF TEST: 04/15/2016 20:37:10

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 20:30 |
| Air Blank          | 0.000 | 20:30 |
| Control Test       | 0.078 | 20:31 |
| Air Blank          | 0.000 | 20:31 |
| Subject Sample # 1 | 0.103 | 20:32 |
| Breath Volume      | 1.898 | 20:32 |
| Air Blank          | 0.000 | 20:33 |
| Air Blank          | 0.000 | 20:34 |
| Subject Sample # 2 | 0.100 | 20:35 |
| Breath Volume      | 2.191 | 20:35 |
| Air Blank          | 0.000 | 20:36 |
| Control Test       | 0.078 | 20:36 |
| Air Blank          | 0.000 | 20:37 |
| Diagnostics Check  | OK    | 20:37 |

---

Data Download Date/Time: 04/30/2016 20:11



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: VILCHES,KEVIN,O  
BREATH TEST OPERATOR AGENCY: Lake County SO  
ARRESTING OFFICER: BLANTON,DA'RRELL  
ARRESTING OFFICER AGENCY: Lake County SO  
LAST AGENCY INSPECTION: 03/31/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1766638  
DRY GAS EXPIRY DATE: 11/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 03:22  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: CUBBERLY

DATE/TIME OF TEST: 04/16/2016 03:47:17

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:40 |
| Air Blank          | 0.000 | 03:41 |
| Control Test       | 0.078 | 03:41 |
| Air Blank          | 0.000 | 03:42 |
| Subject Sample # 1 | 0.000 | 03:42 |
| Breath Volume      | 1.332 | 03:42 |
| Air Blank          | 0.000 | 03:43 |
| Air Blank          | 0.000 | 03:45 |
| Subject Sample # 2 | 0.000 | 03:45 |
| Breath Volume      | 3.257 | 03:45 |
| Air Blank          | 0.000 | 03:46 |
| Control Test       | 0.078 | 03:46 |
| Air Blank          | 0.000 | 03:47 |
| Diagnostics Check  | OK    | 03:47 |

---

Data Download Date/Time: 04/30/2016 20:11

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BELL,ETHAN,C  
BREATH TEST OPERATOR AGENCY: Clermont PD  
ARRESTING OFFICER: BELL,ETHAN  
ARRESTING OFFICER AGENCY: Clermont PD  
LAST AGENCY INSPECTION: 03/31/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1766638  
DRY GAS EXPIRY DATE: 11/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 22:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:24

SUBJECT'S LAST NAME: VIDAL

DATE/TIME OF TEST: 04/27/2016 22:51:21

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:44 |
| Air Blank          | 0.000 | 22:45 |
| Control Test       | 0.079 | 22:45 |
| Air Blank          | 0.000 | 22:46 |
| Subject Sample # 1 | 0.246 | 22:46 |
| Breath Volume      | 1.687 | 22:46 |
| Air Blank          | 0.000 | 22:47 |
| Air Blank          | 0.000 | 22:48 |
| Subject Sample # 2 | 0.245 | 22:49 |
| Breath Volume      | 1.898 | 22:49 |
| Air Blank          | 0.000 | 22:50 |
| Control Test       | 0.078 | 22:50 |
| Air Blank          | 0.000 | 22:51 |
| Diagnostics Check  | OK    | 22:51 |

---

Data Download Date/Time: 04/30/2016 20:11

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FESTA,J,  
BREATH TEST OPERATOR AGENCY: Groveland PD  
ARRESTING OFFICER: FESTA,J  
ARRESTING OFFICER AGENCY: Groveland PD  
LAST AGENCY INSPECTION: 03/31/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1766638  
DRY GAS EXPIRY DATE: 11/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 22:27  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: BAIJNATH

DATE/TIME OF TEST: 04/30/2016 00:25:24

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 00:22 |                      |
| Air Blank         | 0.000 | 00:22 |                      |
| Control Test      | 0.079 | 00:22 |                      |
| Air Blank         | 0.000 | 00:23 |                      |
| Subject Sample    | # 1   | 00:24 | Subject Test Refused |
| Breath Volume     | 0.000 | 00:24 |                      |
| Air Blank         | 0.000 | 00:24 |                      |
| Control Test      | 0.079 | 00:24 |                      |
| Air Blank         | 0.000 | 00:25 |                      |
| Diagnostics Check | OK    | 00:25 |                      |

---

Data Download Date/Time: 04/30/2016 20:11

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000825

INSTRUMENT REGISTERED TO: MOUNT DORA P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RICE,ANDREW,J  
BREATH TEST OPERATOR AGENCY: Mount Dora PD  
ARRESTING OFFICER: DEERING,CONOR  
ARRESTING OFFICER AGENCY: Mount Dora PD  
LAST AGENCY INSPECTION: 04/23/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1841081  
DRY GAS EXPIRY DATE: 04/08/2018  
VIOLATION CODE: DUI  
ARREST TIME: 19:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:10

SUBJECT'S LAST NAME: BUNDSCHUH

DATE/TIME OF TEST: 04/24/2016 20:40:37

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 20:34 |
| Air Blank          | 0.000 | 20:34 |
| Control Test       | 0.078 | 20:35 |
| Air Blank          | 0.000 | 20:35 |
| Subject Sample # 1 | 0.212 | 20:36 |
| Breath Volume      | 3.332 | 20:36 |
| Air Blank          | 0.000 | 20:36 |
| Air Blank          | 0.000 | 20:38 |
| Subject Sample # 2 | 0.213 | 20:39 |
| Breath Volume      | 3.152 | 20:39 |
| Air Blank          | 0.000 | 20:39 |
| Control Test       | 0.080 | 20:40 |
| Air Blank          | 0.000 | 20:40 |
| Diagnostics Check  | OK    | 20:40 |

---

Data Download Date/Time: 05/31/2016 04:32

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DOBOSIEWICZ, ANTHONY, N  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: DOBOSIEWICZ, ANTHONY  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 04/28/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 83550  
DRY GAS EXPIRY DATE: 09/26/2017  
VIOLATION CODE: DUI  
ARREST TIME: 03:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: MONCADA

DATE/TIME OF TEST: 04/29/2016 04:31:52

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:25 |
| Air Blank          | 0.000 | 04:25 |
| Control Test       | 0.080 | 04:25 |
| Air Blank          | 0.000 | 04:26 |
| Subject Sample # 1 | 0.150 | 04:27 |
| Breath Volume      | 1.468 | 04:27 |
| Air Blank          | 0.000 | 04:27 |
| Air Blank          | 0.000 | 04:29 |
| Subject Sample # 2 | 0.143 | 04:30 |
| Breath Volume      | 1.167 | 04:30 |
| Air Blank          | 0.000 | 04:30 |
| Control Test       | 0.080 | 04:31 |
| Air Blank          | 0.000 | 04:31 |
| Diagnostics Check  | OK    | 04:31 |

---

Data Download Date/Time: 05/27/2016 12:14

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MILLER,GARY,A  
BREATH TEST OPERATOR AGENCY: Marion County SO  
ARRESTING OFFICER: MILLER,GARY  
ARRESTING OFFICER AGENCY: Marion County SO  
LAST AGENCY INSPECTION: 04/28/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 83550  
DRY GAS EXPIRY DATE: 09/26/2017  
VIOLATION CODE: DUI  
ARREST TIME: 02:23  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: GROEB

DATE/TIME OF TEST: 04/30/2016 03:32:22

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:26 |
| Air Blank          | 0.000 | 03:26 |
| Control Test       | 0.079 | 03:26 |
| Air Blank          | 0.000 | 03:27 |
| Subject Sample # 1 | 0.098 | 03:27 |
| Breath Volume      | 2.902 | 03:27 |
| Air Blank          | 0.000 | 03:28 |
| Air Blank          | 0.000 | 03:30 |
| Subject Sample # 2 | 0.091 | 03:30 |
| Breath Volume      | 2.281 | 03:30 |
| Air Blank          | 0.000 | 03:31 |
| Control Test       | 0.080 | 03:31 |
| Air Blank          | 0.000 | 03:32 |
| Diagnostics Check  | OK    | 03:32 |

---

Data Download Date/Time: 05/27/2016 12:14

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MCMINN,KATHY,  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: DAWSY,B  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18015080A2  
DRY GAS EXPIRY DATE: 08/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 23:56  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:29

SUBJECT'S LAST NAME: POPP

DATE/TIME OF TEST: 04/07/2016 00:55:41

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:49 |
| Air Blank          | 0.000 | 00:49 |
| Control Test       | 0.078 | 00:50 |
| Air Blank          | 0.000 | 00:50 |
| Subject Sample # 1 | 0.119 | 00:51 |
| Breath Volume      | 1.878 | 00:51 |
| Air Blank          | 0.000 | 00:51 |
| Air Blank          | 0.000 | 00:53 |
| Subject Sample # 2 | 0.110 | 00:54 |
| Breath Volume      | 1.160 | 00:54 |
| Air Blank          | 0.000 | 00:54 |
| Control Test       | 0.079 | 00:55 |
| Air Blank          | 0.000 | 00:55 |
| Diagnostics Check  | OK    | 00:55 |

---

Data Download Date/Time: 05/11/2016 11:48

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BLANCHARD,J  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18015080A2  
DRY GAS EXPIRY DATE: 08/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 00:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:26

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 04/07/2016 01:58:37

---

|                   |       |       |                    |
|-------------------|-------|-------|--------------------|
| Diagnostics Check | OK    | 01:47 |                    |
| Air Blank         | 0.000 | 01:47 |                    |
| Control Test      | 0.079 | 01:48 |                    |
| Air Blank         | 0.000 | 01:48 |                    |
| Subject Sample    | # 1   | 01:51 | No Sample Provided |
| Breath Volume     | 0.000 | 01:51 |                    |
| Air Blank         | 0.000 | 01:52 |                    |
| Air Blank         | 0.000 | 01:54 |                    |
| Subject Sample    | # 2   | 01:57 | No Sample Provided |
| Breath Volume     | 0.000 | 01:57 |                    |
| Air Blank         | 0.000 | 01:57 |                    |
| Control Test      | 0.080 | 01:58 |                    |
| Air Blank         | 0.000 | 01:58 |                    |
| Diagnostics Check | OK    | 01:58 |                    |

---

Data Download Date/Time: 05/11/2016 11:49



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: DE JESUS,V  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18015080A2  
DRY GAS EXPIRY DATE: 08/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 05:17  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 05:56

SUBJECT'S LAST NAME: BRUNK

DATE/TIME OF TEST: 04/07/2016 06:23:23

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 06:16 |
| Air Blank          | 0.000 | 06:17 |
| Control Test       | 0.079 | 06:17 |
| Air Blank          | 0.000 | 06:18 |
| Subject Sample # 1 | 0.205 | 06:18 |
| Breath Volume      | 1.128 | 06:18 |
| Air Blank          | 0.000 | 06:19 |
| Air Blank          | 0.000 | 06:21 |
| Subject Sample # 2 | 0.220 | 06:21 |
| Breath Volume      | 2.007 | 06:21 |
| Air Blank          | 0.000 | 06:22 |
| Control Test       | 0.079 | 06:22 |
| Air Blank          | 0.000 | 06:23 |
| Diagnostics Check  | OK    | 06:23 |

---

Data Download Date/Time: 05/11/2016 11:49

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: FRANKLIN, R  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18015080A2  
DRY GAS EXPIRY DATE: 08/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 20:56  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:43

SUBJECT'S LAST NAME: MORALES CHAVEZ

DATE/TIME OF TEST: 04/07/2016 22:11:30

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:03 |
| Air Blank          | 0.000 | 22:04 |
| Control Test       | 0.079 | 22:04 |
| Air Blank          | 0.000 | 22:05 |
| Subject Sample # 1 | 0.214 | 22:06 |
| Breath Volume      | 1.117 | 22:06 |
| Air Blank          | 0.000 | 22:07 |
| Air Blank          | 0.000 | 22:09 |
| Subject Sample # 2 | 0.203 | 22:10 |
| Breath Volume      | 1.625 | 22:10 |
| Air Blank          | 0.000 | 22:10 |
| Control Test       | 0.077 | 22:10 |
| Air Blank          | 0.000 | 22:11 |
| Diagnostics Check  | OK    | 22:11 |

---

Data Download Date/Time: 05/11/2016 11:49

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: VAN-TREESE, S  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18015080A2  
DRY GAS EXPIRY DATE: 08/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 21:52  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:24

SUBJECT'S LAST NAME: KEPLER

DATE/TIME OF TEST: 04/08/2016 22:51:11

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:44 |
| Air Blank          | 0.000 | 22:44 |
| Control Test       | 0.079 | 22:45 |
| Air Blank          | 0.000 | 22:45 |
| Subject Sample # 1 | 0.040 | 22:46 |
| Breath Volume      | 1.519 | 22:46 |
| Air Blank          | 0.000 | 22:47 |
| Air Blank          | 0.000 | 22:49 |
| Subject Sample # 2 | 0.036 | 22:49 |
| Breath Volume      | 2.468 | 22:49 |
| Air Blank          | 0.000 | 22:50 |
| Control Test       | 0.079 | 22:50 |
| Air Blank          | 0.000 | 22:51 |
| Diagnostics Check  | OK    | 22:51 |

---

Data Download Date/Time: 05/11/2016 11:49

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SARRO,TAYLOR,C  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BRABAND,M  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18015080A2  
DRY GAS EXPIRY DATE: 08/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 23:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:08

SUBJECT'S LAST NAME: MCWILLIAMS

DATE/TIME OF TEST: 04/09/2016 00:35:33

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:28 |
| Air Blank          | 0.000 | 00:29 |
| Control Test       | 0.079 | 00:29 |
| Air Blank          | 0.000 | 00:30 |
| Subject Sample # 1 | 0.224 | 00:31 |
| Breath Volume      | 2.175 | 00:31 |
| Air Blank          | 0.000 | 00:31 |
| Air Blank          | 0.000 | 00:33 |
| Subject Sample # 2 | 0.213 | 00:34 |
| Breath Volume      | 1.332 | 00:34 |
| Air Blank          | 0.000 | 00:34 |
| Control Test       | 0.078 | 00:34 |
| Air Blank          | 0.000 | 00:35 |
| Diagnostics Check  | OK    | 00:35 |

---

Data Download Date/Time: 05/11/2016 11:49

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HENRY,PATRICIA,  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: SHEA,D  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18015080A2  
DRY GAS EXPIRY DATE: 08/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 01:06  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: NATION

DATE/TIME OF TEST: 04/09/2016 01:59:25

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:53 |
| Air Blank          | 0.000 | 01:53 |
| Control Test       | 0.080 | 01:53 |
| Air Blank          | 0.000 | 01:54 |
| Subject Sample # 1 | 0.225 | 01:54 |
| Breath Volume      | 2.449 | 01:54 |
| Air Blank          | 0.000 | 01:55 |
| Air Blank          | 0.000 | 01:57 |
| Subject Sample # 2 | 0.223 | 01:57 |
| Breath Volume      | 2.148 | 01:57 |
| Air Blank          | 0.000 | 01:58 |
| Control Test       | 0.079 | 01:58 |
| Air Blank          | 0.000 | 01:59 |
| Diagnostics Check  | OK    | 01:59 |

---

Data Download Date/Time: 05/11/2016 11:49

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TORRES,JENIFER,  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: HOBSON,D  
ARRESTING OFFICER AGENCY: U of So FL Campus PD  
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18015080A2  
DRY GAS EXPIRY DATE: 08/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 02:05  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:48

SUBJECT'S LAST NAME: HAMMILL

DATE/TIME OF TEST: 04/09/2016 03:24:00

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:17 |
| Air Blank          | 0.000 | 03:18 |
| Control Test       | 0.079 | 03:18 |
| Air Blank          | 0.000 | 03:18 |
| Subject Sample # 1 | 0.205 | 03:19 |
| Breath Volume      | 2.269 | 03:19 |
| Air Blank          | 0.000 | 03:19 |
| Air Blank          | 0.000 | 03:21 |
| Subject Sample # 2 | 0.215 | 03:22 |
| Breath Volume      | 2.964 | 03:22 |
| Air Blank          | 0.000 | 03:23 |
| Control Test       | 0.078 | 03:23 |
| Air Blank          | 0.000 | 03:23 |
| Diagnostics Check  | OK    | 03:24 |

---

Data Download Date/Time: 05/11/2016 11:49

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: WOODEN, D  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18015080A2  
DRY GAS EXPIRY DATE: 08/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 04:07  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: ANTONIO ANTONIO

DATE/TIME OF TEST: 04/09/2016 04:47:08

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:40 |
| Air Blank          | 0.000 | 04:40 |
| Control Test       | 0.079 | 04:41 |
| Air Blank          | 0.000 | 04:41 |
| Subject Sample # 1 | 0.145 | 04:42 |
| Breath Volume      | 1.921 | 04:42 |
| Air Blank          | 0.000 | 04:43 |
| Air Blank          | 0.000 | 04:44 |
| Subject Sample # 2 | 0.150 | 04:45 |
| Breath Volume      | 1.320 | 04:45 |
| Air Blank          | 0.000 | 04:46 |
| Control Test       | 0.078 | 04:46 |
| Air Blank          | 0.000 | 04:47 |
| Diagnostics Check  | OK    | 04:47 |

---

Data Download Date/Time: 05/11/2016 11:49

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HENRY,PATRICIA,  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: LYON,M  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18015080A2  
DRY GAS EXPIRY DATE: 08/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 04:11  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:37

SUBJECT'S LAST NAME: OCHOA

DATE/TIME OF TEST: 04/09/2016 05:08:55

---

|                   |       |       |                    |
|-------------------|-------|-------|--------------------|
| Diagnostics Check | OK    | 04:57 |                    |
| Air Blank         | 0.000 | 04:57 |                    |
| Control Test      | 0.079 | 04:58 |                    |
| Air Blank         | 0.000 | 04:58 |                    |
| Subject Sample    | # 1   | 05:01 | No Sample Provided |
| Breath Volume     | 0.000 | 05:01 |                    |
| Air Blank         | 0.000 | 05:02 |                    |
| Air Blank         | 0.000 | 05:04 |                    |
| Subject Sample    | # 2   | 05:07 | Volume Not Met     |
| Breath Volume     | 0.058 | 05:07 |                    |
| Air Blank         | 0.000 | 05:08 |                    |
| Control Test      | 0.078 | 05:08 |                    |
| Air Blank         | 0.000 | 05:08 |                    |
| Diagnostics Check | OK    | 05:08 |                    |

---

Data Download Date/Time: 05/11/2016 11:49



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: PINEO,L  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18015080A2  
DRY GAS EXPIRY DATE: 08/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 18:04  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 19:11

SUBJECT'S LAST NAME: DAVISON

DATE/TIME OF TEST: 04/09/2016 19:39:27

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 19:32 |
| Air Blank          | 0.000 | 19:33 |
| Control Test       | 0.079 | 19:33 |
| Air Blank          | 0.000 | 19:34 |
| Subject Sample # 1 | 0.254 | 19:34 |
| Breath Volume      | 1.976 | 19:34 |
| Air Blank          | 0.000 | 19:35 |
| Air Blank          | 0.000 | 19:37 |
| Subject Sample # 2 | 0.245 | 19:37 |
| Breath Volume      | 1.953 | 19:37 |
| Air Blank          | 0.000 | 19:38 |
| Control Test       | 0.080 | 19:38 |
| Air Blank          | 0.000 | 19:39 |
| Diagnostics Check  | OK    | 19:39 |

---

Data Download Date/Time: 05/11/2016 11:49

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MCMINN,KATHY,  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: GARRELL,Q  
ARRESTING OFFICER AGENCY: Plant City PD  
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18015080A2  
DRY GAS EXPIRY DATE: 08/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 22:16  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:04

SUBJECT'S LAST NAME: LOPEZ-PEDRO

DATE/TIME OF TEST: 04/09/2016 23:30:59

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:24 |
| Air Blank          | 0.000 | 23:25 |
| Control Test       | 0.080 | 23:25 |
| Air Blank          | 0.000 | 23:25 |
| Subject Sample # 1 | 0.176 | 23:26 |
| Breath Volume      | 2.277 | 23:26 |
| Air Blank          | 0.000 | 23:26 |
| Air Blank          | 0.000 | 23:28 |
| Subject Sample # 2 | 0.182 | 23:29 |
| Breath Volume      | 1.828 | 23:29 |
| Air Blank          | 0.000 | 23:30 |
| Control Test       | 0.080 | 23:30 |
| Air Blank          | 0.000 | 23:30 |
| Diagnostics Check  | OK    | 23:30 |

---

Data Download Date/Time: 05/11/2016 11:49

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KIRK,CASEY,S  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: PALFY,R  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18015080A2  
DRY GAS EXPIRY DATE: 08/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 03:18  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: CHAVEZ

DATE/TIME OF TEST: 04/10/2016 04:16:21

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:10 |
| Air Blank          | 0.000 | 04:10 |
| Control Test       | 0.079 | 04:11 |
| Air Blank          | 0.000 | 04:11 |
| Subject Sample # 1 | 0.082 | 04:12 |
| Breath Volume      | 1.980 | 04:12 |
| Air Blank          | 0.000 | 04:12 |
| Air Blank          | 0.000 | 04:14 |
| Subject Sample # 2 | 0.082 | 04:14 |
| Breath Volume      | 2.089 | 04:14 |
| Air Blank          | 0.000 | 04:15 |
| Control Test       | 0.079 | 04:15 |
| Air Blank          | 0.000 | 04:16 |
| Diagnostics Check  | OK    | 04:16 |

---

Data Download Date/Time: 05/11/2016 11:49

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MCMINN,KATHY,  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: MALONE,C  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18015080A2  
DRY GAS EXPIRY DATE: 08/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 03:29  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:11

SUBJECT'S LAST NAME: TREJO

DATE/TIME OF TEST: 04/10/2016 04:38:23

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:31 |
| Air Blank          | 0.000 | 04:32 |
| Control Test       | 0.079 | 04:32 |
| Air Blank          | 0.000 | 04:32 |
| Subject Sample # 1 | 0.125 | 04:33 |
| Breath Volume      | 1.218 | 04:33 |
| Air Blank          | 0.000 | 04:34 |
| Air Blank          | 0.000 | 04:36 |
| Subject Sample # 2 | 0.122 | 04:36 |
| Breath Volume      | 1.210 | 04:36 |
| Air Blank          | 0.000 | 04:37 |
| Control Test       | 0.079 | 04:37 |
| Air Blank          | 0.000 | 04:38 |
| Diagnostics Check  | OK    | 04:38 |

---

Data Download Date/Time: 05/11/2016 11:49

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SMITH,REBECCA,J  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BRABAND,M  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18015080A2  
DRY GAS EXPIRY DATE: 08/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 05:58  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:44

SUBJECT'S LAST NAME: MORENCY

DATE/TIME OF TEST: 04/10/2016 07:19:46

---

|                   |           |       |                |
|-------------------|-----------|-------|----------------|
| Diagnostics Check | OK        | 07:06 |                |
| Air Blank         | 0.000     | 07:07 |                |
| Control Test      | 0.080     | 07:07 |                |
| Air Blank         | 0.000     | 07:08 |                |
| Subject Sample    | # 1 0.113 | 07:11 | Volume Not Met |
| Breath Volume     | 0.882     | 07:11 |                |
| Air Blank         | 0.000     | 07:11 |                |
| Air Blank         | 0.000     | 07:13 |                |
| Subject Sample    | # 2 0.142 | 07:14 |                |
| Breath Volume     | 2.496     | 07:14 |                |
| Air Blank         | 0.000     | 07:15 |                |
| Air Blank         | 0.000     | 07:17 |                |
| Subject Sample    | # 3 0.140 | 07:18 |                |
| Breath Volume     | 2.296     | 07:18 |                |
| Air Blank         | 0.000     | 07:18 |                |
| Control Test      | 0.079     | 07:19 |                |
| Air Blank         | 0.000     | 07:19 |                |
| Diagnostics Check | OK        | 07:19 |                |

---

Data Download Date/Time: 05/11/2016 11:49

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SMITH,REBECCA,J  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: DEVORE,J  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18015080A2  
DRY GAS EXPIRY DATE: 08/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 07:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 07:45

SUBJECT'S LAST NAME: MOON

DATE/TIME OF TEST: 04/10/2016 08:19:12

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 08:08 |
| Air Blank          | 0.000 | 08:08 |
| Control Test       | 0.079 | 08:09 |
| Air Blank          | 0.000 | 08:09 |
| Subject Sample # 1 | 0.119 | 08:12 |
| Breath Volume      | 1.195 | 08:12 |
| Air Blank          | 0.000 | 08:13 |
| Air Blank          | 0.000 | 08:15 |
| Subject Sample # 2 | 0.121 | 08:17 |
| Breath Volume      | 1.230 | 08:17 |
| Air Blank          | 0.000 | 08:18 |
| Control Test       | 0.079 | 08:18 |
| Air Blank          | 0.000 | 08:19 |
| Diagnostics Check  | OK    | 08:19 |

---

Data Download Date/Time: 05/11/2016 11:49

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KIRK,CASEY,S  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: VAN-TREESE,S  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18015080A2  
DRY GAS EXPIRY DATE: 08/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 00:57  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:21

SUBJECT'S LAST NAME: MCBRIDE

DATE/TIME OF TEST: 04/11/2016 01:47:34

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:41 |
| Air Blank          | 0.000 | 01:41 |
| Control Test       | 0.080 | 01:42 |
| Air Blank          | 0.000 | 01:42 |
| Subject Sample # 1 | 0.151 | 01:43 |
| Breath Volume      | 1.675 | 01:43 |
| Air Blank          | 0.000 | 01:43 |
| Air Blank          | 0.000 | 01:45 |
| Subject Sample # 2 | 0.148 | 01:46 |
| Breath Volume      | 1.339 | 01:46 |
| Air Blank          | 0.000 | 01:46 |
| Control Test       | 0.079 | 01:46 |
| Air Blank          | 0.000 | 01:47 |
| Diagnostics Check  | OK    | 01:47 |

---

Data Download Date/Time: 05/11/2016 11:49

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MCMINN,KATHY,  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: MALONE,C  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18015080A2  
DRY GAS EXPIRY DATE: 08/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 01:29  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: ALI

DATE/TIME OF TEST: 04/11/2016 02:33:16

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:26 |
| Air Blank          | 0.000 | 02:27 |
| Control Test       | 0.079 | 02:27 |
| Air Blank          | 0.000 | 02:28 |
| Subject Sample # 1 | 0.254 | 02:28 |
| Breath Volume      | 1.105 | 02:28 |
| Air Blank          | 0.000 | 02:29 |
| Air Blank          | 0.000 | 02:31 |
| Subject Sample # 2 | 0.247 | 02:31 |
| Breath Volume      | 1.101 | 02:31 |
| Air Blank          | 0.000 | 02:32 |
| Control Test       | 0.076 | 02:32 |
| Air Blank          | 0.000 | 02:33 |
| Diagnostics Check  | OK    | 02:33 |

---

Data Download Date/Time: 05/11/2016 11:49



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KIRK,CASEY,S  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: DEJESUS,V  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18015080A2  
DRY GAS EXPIRY DATE: 08/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 00:43  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 04/12/2016 02:05:41

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 02:00 |                      |
| Air Blank         | 0.000 | 02:00 |                      |
| Control Test      | 0.079 | 02:01 |                      |
| Air Blank         | 0.000 | 02:01 |                      |
| Subject Sample    | # 1   | 02:04 | Subject Test Refused |
| Breath Volume     | 0.000 | 02:04 |                      |
| Air Blank         | 0.000 | 02:04 |                      |
| Control Test      | 0.079 | 02:05 |                      |
| Air Blank         | 0.000 | 02:05 |                      |
| Diagnostics Check | OK    | 02:05 |                      |

---

Data Download Date/Time: 05/11/2016 11:49

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BARLAUG,G  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18015080A2  
DRY GAS EXPIRY DATE: 08/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 02:17  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:47

SUBJECT'S LAST NAME: REILLY

DATE/TIME OF TEST: 04/12/2016 03:19:07

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:07 |
| Air Blank          | 0.000 | 03:07 |
| Control Test       | 0.079 | 03:08 |
| Air Blank          | 0.000 | 03:08 |
| Subject Sample # 1 | 0.242 | 03:10 |
| Breath Volume      | 2.894 | 03:10 |
| Air Blank          | 0.000 | 03:11 |
| Air Blank          | 0.000 | 03:13 |
| Subject Sample # 2 | 0.267 | 03:13 |
| Breath Volume      | 2.238 | 03:13 |
| Air Blank          | 0.000 | 03:14 |
| Air Blank          | 0.000 | 03:16 |
| Subject Sample # 3 | 0.251 | 03:17 |
| Breath Volume      | 1.371 | 03:17 |
| Air Blank          | 0.000 | 03:18 |
| Control Test       | 0.078 | 03:18 |
| Air Blank          | 0.000 | 03:19 |
| Diagnostics Check  | OK    | 03:19 |

---

Data Download Date/Time: 05/11/2016 11:49

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TORRES,JENIFER,  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: DEJESUS,V  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18015080A2  
DRY GAS EXPIRY DATE: 08/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 18:26  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:57

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 04/12/2016 19:25:42

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 19:18 |
| Air Blank          | 0.000 | 19:18 |
| Control Test       | 0.079 | 19:19 |
| Air Blank          | 0.000 | 19:19 |
| Subject Sample # 1 | 0.112 | 19:21 |
| Breath Volume      | 1.566 | 19:21 |
| Air Blank          | 0.000 | 19:21 |
| Air Blank          | 0.000 | 19:23 |
| Subject Sample # 2 | 0.108 | 19:24 |
| Breath Volume      | 1.382 | 19:24 |
| Air Blank          | 0.000 | 19:24 |
| Control Test       | 0.079 | 19:25 |
| Air Blank          | 0.000 | 19:25 |
| Diagnostics Check  | OK    | 19:25 |

---

Data Download Date/Time: 05/11/2016 11:49

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: VISSER, A  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18015080A2  
DRY GAS EXPIRY DATE: 08/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 03:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:39

SUBJECT'S LAST NAME: LENNON

DATE/TIME OF TEST: 04/13/2016 04:06:17

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:00 |
| Air Blank          | 0.000 | 04:00 |
| Control Test       | 0.079 | 04:00 |
| Air Blank          | 0.000 | 04:01 |
| Subject Sample # 1 | 0.167 | 04:01 |
| Breath Volume      | 1.695 | 04:01 |
| Air Blank          | 0.000 | 04:02 |
| Air Blank          | 0.000 | 04:04 |
| Subject Sample # 2 | 0.184 | 04:04 |
| Breath Volume      | 1.148 | 04:04 |
| Air Blank          | 0.000 | 04:05 |
| Control Test       | 0.078 | 04:05 |
| Air Blank          | 0.000 | 04:06 |
| Diagnostics Check  | OK    | 04:06 |

---

Data Download Date/Time: 05/11/2016 11:49

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: NIEUWENDAM, R  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18015080A2  
DRY GAS EXPIRY DATE: 08/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 20:07  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:04

SUBJECT'S LAST NAME: PITCHER

DATE/TIME OF TEST: 04/13/2016 21:35:39

---

|                   |       |       |                    |
|-------------------|-------|-------|--------------------|
| Diagnostics Check | OK    | 21:24 |                    |
| Air Blank         | 0.000 | 21:24 |                    |
| Control Test      | 0.079 | 21:24 |                    |
| Air Blank         | 0.000 | 21:25 |                    |
| Subject Sample    | # 1   | 21:28 | No Sample Provided |
| Breath Volume     | 0.000 | 21:28 |                    |
| Air Blank         | 0.000 | 21:28 |                    |
| Air Blank         | 0.000 | 21:31 |                    |
| Subject Sample    | # 2   | 21:34 | Volume Not Met     |
| Breath Volume     | 0.238 | 21:34 |                    |
| Air Blank         | 0.000 | 21:34 |                    |
| Control Test      | 0.078 | 21:35 |                    |
| Air Blank         | 0.000 | 21:35 |                    |
| Diagnostics Check | OK    | 21:35 |                    |

---

Data Download Date/Time: 05/11/2016 11:49

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: SIMPKINS, D  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18015080A2  
DRY GAS EXPIRY DATE: 08/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 01:15  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:52

SUBJECT'S LAST NAME: MITCHELL

DATE/TIME OF TEST: 04/14/2016 02:21:25

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:13 |
| Air Blank          | 0.000 | 02:13 |
| Control Test       | 0.079 | 02:13 |
| Air Blank          | 0.000 | 02:14 |
| Subject Sample # 1 | 0.000 | 02:15 |
| Breath Volume      | 1.261 | 02:15 |
| Air Blank          | 0.000 | 02:16 |
| Air Blank          | 0.000 | 02:18 |
| Subject Sample # 2 | 0.000 | 02:20 |
| Breath Volume      | 1.261 | 02:20 |
| Air Blank          | 0.000 | 02:20 |
| Control Test       | 0.080 | 02:20 |
| Air Blank          | 0.000 | 02:21 |
| Diagnostics Check  | OK    | 02:21 |

---

Data Download Date/Time: 05/11/2016 11:49

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TORRES,JENIFER,  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: MANGANARO,J  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18015080A2  
DRY GAS EXPIRY DATE: 08/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 03:15  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:04

SUBJECT'S LAST NAME: GOLDNER

DATE/TIME OF TEST: 04/14/2016 04:39:33

---

|                   |           |       |                |
|-------------------|-----------|-------|----------------|
| Diagnostics Check | OK        | 04:25 |                |
| Air Blank         | 0.000     | 04:25 |                |
| Control Test      | 0.080     | 04:26 |                |
| Air Blank         | 0.000     | 04:26 |                |
| Subject Sample    | # 1 0.066 | 04:29 | Volume Not Met |
| Breath Volume     | 1.292     | 04:29 |                |
| Air Blank         | 0.000     | 04:30 |                |
| Air Blank         | 0.000     | 04:32 |                |
| Subject Sample    | # 2 0.072 | 04:33 |                |
| Breath Volume     | 1.683     | 04:33 |                |
| Air Blank         | 0.000     | 04:34 |                |
| Air Blank         | 0.000     | 04:36 |                |
| Subject Sample    | # 3 0.072 | 04:38 |                |
| Breath Volume     | 1.195     | 04:38 |                |
| Air Blank         | 0.000     | 04:38 |                |
| Control Test      | 0.079     | 04:38 |                |
| Air Blank         | 0.000     | 04:39 |                |
| Diagnostics Check | OK        | 04:39 |                |

---

Data Download Date/Time: 05/11/2016 11:49

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SARRO,TAYLOR,C  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: ANGELICCHI,J  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18015080A2  
DRY GAS EXPIRY DATE: 08/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 04:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:18

SUBJECT'S LAST NAME: BAXLEY JR

DATE/TIME OF TEST: 04/14/2016 05:45:08

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 05:39 |
| Air Blank          | 0.000 | 05:39 |
| Control Test       | 0.079 | 05:39 |
| Air Blank          | 0.000 | 05:40 |
| Subject Sample # 1 | 0.179 | 05:40 |
| Breath Volume      | 2.007 | 05:40 |
| Air Blank          | 0.000 | 05:41 |
| Air Blank          | 0.000 | 05:43 |
| Subject Sample # 2 | 0.180 | 05:43 |
| Breath Volume      | 2.082 | 05:43 |
| Air Blank          | 0.000 | 05:44 |
| Control Test       | 0.079 | 05:44 |
| Air Blank          | 0.000 | 05:45 |
| Diagnostics Check  | OK    | 05:45 |

---

Data Download Date/Time: 05/11/2016 11:49



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KIRK,CASEY,S  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: SNAGGS,S  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18015080A2  
DRY GAS EXPIRY DATE: 08/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 02:01  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:46

SUBJECT'S LAST NAME: MARCANO

DATE/TIME OF TEST: 04/15/2016 03:13:41

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:06 |
| Air Blank          | 0.000 | 03:07 |
| Control Test       | 0.079 | 03:07 |
| Air Blank          | 0.000 | 03:08 |
| Subject Sample # 1 | 0.095 | 03:08 |
| Breath Volume      | 1.128 | 03:08 |
| Air Blank          | 0.000 | 03:09 |
| Air Blank          | 0.000 | 03:11 |
| Subject Sample # 2 | 0.093 | 03:12 |
| Breath Volume      | 1.562 | 03:12 |
| Air Blank          | 0.000 | 03:12 |
| Control Test       | 0.079 | 03:13 |
| Air Blank          | 0.000 | 03:13 |
| Diagnostics Check  | OK    | 03:13 |

---

Data Download Date/Time: 05/11/2016 11:49

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KIRK,CASEY,S  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: HERNANDEZ,A  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18015080A2  
DRY GAS EXPIRY DATE: 08/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 02:03  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:46

SUBJECT'S LAST NAME: TOMAS-HERRERA

DATE/TIME OF TEST: 04/16/2016 03:13:15

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:07 |
| Air Blank          | 0.000 | 03:07 |
| Control Test       | 0.080 | 03:07 |
| Air Blank          | 0.000 | 03:08 |
| Subject Sample # 1 | 0.160 | 03:08 |
| Breath Volume      | 2.222 | 03:08 |
| Air Blank          | 0.000 | 03:09 |
| Air Blank          | 0.000 | 03:11 |
| Subject Sample # 2 | 0.164 | 03:11 |
| Breath Volume      | 3.664 | 03:11 |
| Air Blank          | 0.000 | 03:12 |
| Control Test       | 0.079 | 03:12 |
| Air Blank          | 0.000 | 03:13 |
| Diagnostics Check  | OK    | 03:13 |

---

Data Download Date/Time: 05/11/2016 11:49

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: DE JESUS,V  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18015080A2  
DRY GAS EXPIRY DATE: 08/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 02:22  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: ANTONIO-DOMINGUEZ

DATE/TIME OF TEST: 04/16/2016 03:36:08

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:28 |
| Air Blank          | 0.000 | 03:29 |
| Control Test       | 0.079 | 03:29 |
| Air Blank          | 0.000 | 03:30 |
| Subject Sample # 1 | 0.197 | 03:31 |
| Breath Volume      | 2.046 | 03:31 |
| Air Blank          | 0.000 | 03:32 |
| Air Blank          | 0.000 | 03:33 |
| Subject Sample # 2 | 0.199 | 03:34 |
| Breath Volume      | 2.265 | 03:34 |
| Air Blank          | 0.000 | 03:35 |
| Control Test       | 0.079 | 03:35 |
| Air Blank          | 0.000 | 03:36 |
| Diagnostics Check  | OK    | 03:36 |

---

Data Download Date/Time: 05/11/2016 11:49

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SARRO,TAYLOR,C  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BRIGHTWELL,R  
ARRESTING OFFICER AGENCY: U of So FL Campus PD  
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18015080A2  
DRY GAS EXPIRY DATE: 08/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 22:56  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: REYES GONZALEZ

DATE/TIME OF TEST: 04/16/2016 23:58:01

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:50 |
| Air Blank          | 0.000 | 23:51 |
| Control Test       | 0.080 | 23:51 |
| Air Blank          | 0.000 | 23:52 |
| Subject Sample # 1 | 0.182 | 23:53 |
| Breath Volume      | 2.843 | 23:53 |
| Air Blank          | 0.000 | 23:54 |
| Air Blank          | 0.000 | 23:56 |
| Subject Sample # 2 | 0.185 | 23:56 |
| Breath Volume      | 1.996 | 23:56 |
| Air Blank          | 0.000 | 23:57 |
| Control Test       | 0.079 | 23:57 |
| Air Blank          | 0.000 | 23:57 |
| Diagnostics Check  | OK    | 23:58 |

---

Data Download Date/Time: 05/11/2016 11:49

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HENRY,PATRICIA,  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: DAWSY,B  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18015080A2  
DRY GAS EXPIRY DATE: 08/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 02:11  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: WELLS SHAHAN

DATE/TIME OF TEST: 04/17/2016 03:02:17

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:55 |
| Air Blank          | 0.000 | 02:56 |
| Control Test       | 0.079 | 02:56 |
| Air Blank          | 0.000 | 02:56 |
| Subject Sample # 1 | 0.138 | 02:57 |
| Breath Volume      | 3.964 | 02:57 |
| Air Blank          | 0.000 | 02:58 |
| Air Blank          | 0.000 | 03:00 |
| Subject Sample # 2 | 0.141 | 03:00 |
| Breath Volume      | 4.277 | 03:00 |
| Air Blank          | 0.000 | 03:01 |
| Control Test       | 0.078 | 03:01 |
| Air Blank          | 0.000 | 03:02 |
| Diagnostics Check  | OK    | 03:02 |

---

Data Download Date/Time: 05/11/2016 11:49

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TORRES,JENIFER,  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BADEN,T  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18015080A2  
DRY GAS EXPIRY DATE: 08/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 04:16  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:31

SUBJECT'S LAST NAME: GRACE

DATE/TIME OF TEST: 04/17/2016 05:01:00

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:54 |
| Air Blank          | 0.000 | 04:54 |
| Control Test       | 0.079 | 04:54 |
| Air Blank          | 0.000 | 04:55 |
| Subject Sample # 1 | 0.107 | 04:56 |
| Breath Volume      | 1.656 | 04:56 |
| Air Blank          | 0.000 | 04:56 |
| Air Blank          | 0.000 | 04:58 |
| Subject Sample # 2 | 0.106 | 04:59 |
| Breath Volume      | 1.753 | 04:59 |
| Air Blank          | 0.000 | 05:00 |
| Control Test       | 0.079 | 05:00 |
| Air Blank          | 0.000 | 05:00 |
| Diagnostics Check  | OK    | 05:01 |

---

Data Download Date/Time: 05/11/2016 11:49

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SARRO,TAYLOR,C  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BAILEY,D  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18015080A2  
DRY GAS EXPIRY DATE: 08/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 23:48  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:24

SUBJECT'S LAST NAME: PEREZ RAMOS

DATE/TIME OF TEST: 04/18/2016 00:51:46

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:44 |
| Air Blank          | 0.000 | 00:44 |
| Control Test       | 0.080 | 00:45 |
| Air Blank          | 0.000 | 00:45 |
| Subject Sample # 1 | 0.176 | 00:46 |
| Breath Volume      | 1.925 | 00:46 |
| Air Blank          | 0.000 | 00:47 |
| Air Blank          | 0.000 | 00:48 |
| Subject Sample # 2 | 0.175 | 00:50 |
| Breath Volume      | 3.484 | 00:50 |
| Air Blank          | 0.000 | 00:50 |
| Control Test       | 0.080 | 00:51 |
| Air Blank          | 0.000 | 00:51 |
| Diagnostics Check  | OK    | 00:51 |

---

Data Download Date/Time: 05/11/2016 11:49

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MCMINN,KATHY,  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: VISSER,A  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18015080A2  
DRY GAS EXPIRY DATE: 08/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 23:31  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:37

SUBJECT'S LAST NAME: MARINO

DATE/TIME OF TEST: 04/21/2016 01:04:56

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:57 |
| Air Blank          | 0.000 | 00:58 |
| Control Test       | 0.079 | 00:58 |
| Air Blank          | 0.000 | 00:59 |
| Subject Sample # 1 | 0.126 | 01:00 |
| Breath Volume      | 1.179 | 01:00 |
| Air Blank          | 0.000 | 01:00 |
| Air Blank          | 0.000 | 01:02 |
| Subject Sample # 2 | 0.124 | 01:03 |
| Breath Volume      | 1.621 | 01:03 |
| Air Blank          | 0.000 | 01:04 |
| Control Test       | 0.078 | 01:04 |
| Air Blank          | 0.000 | 01:04 |
| Diagnostics Check  | OK    | 01:04 |

---

Data Download Date/Time: 05/11/2016 11:49



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KIRK,CASEY,S  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: SNAGGS,S  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18015080A2  
DRY GAS EXPIRY DATE: 08/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 04:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:50

SUBJECT'S LAST NAME: STEEL

DATE/TIME OF TEST: 04/21/2016 05:18:28

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 05:10 |
| Air Blank          | 0.000 | 05:10 |
| Control Test       | 0.079 | 05:10 |
| Air Blank          | 0.000 | 05:11 |
| Subject Sample # 1 | 0.113 | 05:11 |
| Breath Volume      | 1.871 | 05:11 |
| Air Blank          | 0.000 | 05:12 |
| Air Blank          | 0.000 | 05:14 |
| Subject Sample # 2 | 0.111 | 05:16 |
| Breath Volume      | 2.703 | 05:16 |
| Air Blank          | 0.000 | 05:17 |
| Control Test       | 0.079 | 05:17 |
| Air Blank          | 0.000 | 05:18 |
| Diagnostics Check  | OK    | 05:18 |

---

Data Download Date/Time: 05/11/2016 11:49

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MCMINN,KATHY,  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: FRANKLIN,R  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18015080A2  
DRY GAS EXPIRY DATE: 08/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 04:48  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:09

SUBJECT'S LAST NAME: MARTIN

DATE/TIME OF TEST: 04/21/2016 05:36:27

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 05:29 |
| Air Blank          | 0.000 | 05:30 |
| Control Test       | 0.079 | 05:30 |
| Air Blank          | 0.000 | 05:31 |
| Subject Sample # 1 | 0.078 | 05:31 |
| Breath Volume      | 2.542 | 05:31 |
| Air Blank          | 0.000 | 05:32 |
| Air Blank          | 0.000 | 05:34 |
| Subject Sample # 2 | 0.077 | 05:34 |
| Breath Volume      | 2.804 | 05:34 |
| Air Blank          | 0.000 | 05:35 |
| Control Test       | 0.078 | 05:35 |
| Air Blank          | 0.000 | 05:36 |
| Diagnostics Check  | OK    | 05:36 |

---

Data Download Date/Time: 05/11/2016 11:49

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: VISSER, A  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18015080A2  
DRY GAS EXPIRY DATE: 08/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 00:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: NICKELL

DATE/TIME OF TEST: 04/22/2016 01:07:59

---

|                   |           |       |                |
|-------------------|-----------|-------|----------------|
| Diagnostics Check | OK        | 00:55 |                |
| Air Blank         | 0.000     | 00:56 |                |
| Control Test      | 0.079     | 00:56 |                |
| Air Blank         | 0.000     | 00:56 |                |
| Subject Sample    | # 1 0.189 | 00:59 | Volume Not Met |
| Breath Volume     | 0.402     | 00:59 |                |
| Air Blank         | 0.000     | 01:00 |                |
| Air Blank         | 0.000     | 01:02 |                |
| Subject Sample    | # 2 0.177 | 01:03 |                |
| Breath Volume     | 1.378     | 01:03 |                |
| Air Blank         | 0.000     | 01:04 |                |
| Air Blank         | 0.000     | 01:06 |                |
| Subject Sample    | # 3 0.172 | 01:06 |                |
| Breath Volume     | 1.226     | 01:06 |                |
| Air Blank         | 0.000     | 01:07 |                |
| Control Test      | 0.077     | 01:07 |                |
| Air Blank         | 0.000     | 01:07 |                |
| Diagnostics Check | OK        | 01:07 |                |

---

Data Download Date/Time: 05/11/2016 11:49

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HENRY, PATRICIA,  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: SNAGGS, S  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18015080A2  
DRY GAS EXPIRY DATE: 08/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 02:57  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:32

SUBJECT'S LAST NAME: PHILLIPS

DATE/TIME OF TEST: 04/22/2016 03:58:52

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:52 |
| Air Blank          | 0.000 | 03:52 |
| Control Test       | 0.079 | 03:53 |
| Air Blank          | 0.000 | 03:53 |
| Subject Sample # 1 | 0.124 | 03:54 |
| Breath Volume      | 2.222 | 03:54 |
| Air Blank          | 0.000 | 03:54 |
| Air Blank          | 0.000 | 03:56 |
| Subject Sample # 2 | 0.134 | 03:57 |
| Breath Volume      | 2.503 | 03:57 |
| Air Blank          | 0.000 | 03:57 |
| Control Test       | 0.078 | 03:58 |
| Air Blank          | 0.000 | 03:58 |
| Diagnostics Check  | OK    | 03:58 |

---

Data Download Date/Time: 05/11/2016 11:49

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TORRES,JENIFER,  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: VISSER,A  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18015080A2  
DRY GAS EXPIRY DATE: 08/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 00:03  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:29

SUBJECT'S LAST NAME: POWALI

DATE/TIME OF TEST: 04/23/2016 01:02:38

---

|                   |       |       |                    |
|-------------------|-------|-------|--------------------|
| Diagnostics Check | OK    | 00:51 |                    |
| Air Blank         | 0.000 | 00:51 |                    |
| Control Test      | 0.080 | 00:52 |                    |
| Air Blank         | 0.000 | 00:52 |                    |
| Subject Sample    | # 1   | 00:55 | No Sample Provided |
| Breath Volume     | 0.000 | 00:55 |                    |
| Air Blank         | 0.000 | 00:56 |                    |
| Air Blank         | 0.000 | 00:58 |                    |
| Subject Sample    | # 2   | 01:01 | No Sample Provided |
| Breath Volume     | 0.000 | 01:01 |                    |
| Air Blank         | 0.000 | 01:01 |                    |
| Control Test      | 0.080 | 01:02 |                    |
| Air Blank         | 0.000 | 01:02 |                    |
| Diagnostics Check | OK    | 01:02 |                    |

---

Data Download Date/Time: 05/11/2016 11:49

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TORRES,JENIFER,  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: CRAGG,S  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18015080A2  
DRY GAS EXPIRY DATE: 08/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 02:39  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:06

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 04/23/2016 03:34:03

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:26 |
| Air Blank          | 0.000 | 03:27 |
| Control Test       | 0.079 | 03:27 |
| Air Blank          | 0.000 | 03:27 |
| Subject Sample # 1 | 0.082 | 03:28 |
| Breath Volume      | 1.148 | 03:28 |
| Air Blank          | 0.000 | 03:29 |
| Air Blank          | 0.000 | 03:31 |
| Subject Sample # 2 | 0.086 | 03:32 |
| Breath Volume      | 1.917 | 03:32 |
| Air Blank          | 0.000 | 03:33 |
| Control Test       | 0.080 | 03:33 |
| Air Blank          | 0.000 | 03:33 |
| Diagnostics Check  | OK    | 03:34 |

---

Data Download Date/Time: 05/11/2016 11:49

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: LYON, M  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18015080A2  
DRY GAS EXPIRY DATE: 08/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 03:12  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:49

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 04/23/2016 04:16:26

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:09 |
| Air Blank          | 0.000 | 04:10 |
| Control Test       | 0.079 | 04:10 |
| Air Blank          | 0.000 | 04:11 |
| Subject Sample # 1 | 0.104 | 04:11 |
| Breath Volume      | 1.714 | 04:11 |
| Air Blank          | 0.000 | 04:12 |
| Air Blank          | 0.000 | 04:14 |
| Subject Sample # 2 | 0.110 | 04:14 |
| Breath Volume      | 2.511 | 04:14 |
| Air Blank          | 0.000 | 04:15 |
| Control Test       | 0.079 | 04:15 |
| Air Blank          | 0.000 | 04:16 |
| Diagnostics Check  | OK    | 04:16 |

---

Data Download Date/Time: 05/11/2016 11:49

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MCMINN,KATHY,  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: SNAGGS,S  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18015080A2  
DRY GAS EXPIRY DATE: 08/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 21:34  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:12

SUBJECT'S LAST NAME: LAVERTY

DATE/TIME OF TEST: 04/23/2016 22:40:47

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:32 |
| Air Blank          | 0.000 | 22:32 |
| Control Test       | 0.080 | 22:32 |
| Air Blank          | 0.000 | 22:33 |
| Subject Sample # 1 | 0.206 | 22:34 |
| Breath Volume      | 1.281 | 22:34 |
| Air Blank          | 0.000 | 22:34 |
| Air Blank          | 0.000 | 22:36 |
| Subject Sample # 2 | 0.207 | 22:39 |
| Breath Volume      | 1.597 | 22:39 |
| Air Blank          | 0.000 | 22:39 |
| Control Test       | 0.080 | 22:40 |
| Air Blank          | 0.000 | 22:40 |
| Diagnostics Check  | OK    | 22:40 |

---

Data Download Date/Time: 05/11/2016 11:49



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KIRK,CASEY,S  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: DUNCAN,M  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18015080A2  
DRY GAS EXPIRY DATE: 08/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 01:16  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:38

SUBJECT'S LAST NAME: DYE

DATE/TIME OF TEST: 04/24/2016 02:10:14

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:01 |
| Air Blank          | 0.000 | 02:01 |
| Control Test       | 0.079 | 02:01 |
| Air Blank          | 0.000 | 02:02 |
| Subject Sample # 1 | 0.230 | 02:02 |
| Breath Volume      | 1.292 | 02:02 |
| Air Blank          | 0.000 | 02:03 |
| Air Blank          | 0.000 | 02:05 |
| Subject Sample # 2 | 0.254 | 02:05 |
| Breath Volume      | 2.914 | 02:05 |
| Air Blank          | 0.000 | 02:06 |
| Air Blank          | 0.000 | 02:08 |
| Subject Sample # 3 | 0.238 | 02:08 |
| Breath Volume      | 1.824 | 02:08 |
| Air Blank          | 0.000 | 02:09 |
| Control Test       | 0.079 | 02:09 |
| Air Blank          | 0.000 | 02:10 |
| Diagnostics Check  | OK    | 02:10 |

---

Data Download Date/Time: 05/11/2016 11:49

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MCMINN,KATHY,  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: PORTMAN,A  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18015080A2  
DRY GAS EXPIRY DATE: 08/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 03:27  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:53

SUBJECT'S LAST NAME: GUZMAN

DATE/TIME OF TEST: 04/24/2016 04:28:38

---

|                   |           |       |                |
|-------------------|-----------|-------|----------------|
| Diagnostics Check | OK        | 04:17 |                |
| Air Blank         | 0.000     | 04:17 |                |
| Control Test      | 0.078     | 04:17 |                |
| Air Blank         | 0.000     | 04:18 |                |
| Subject Sample    | # 1 0.177 | 04:21 | Volume Not Met |
| Breath Volume     | 0.523     | 04:21 |                |
| Air Blank         | 0.000     | 04:22 |                |
| Air Blank         | 0.000     | 04:23 |                |
| Subject Sample    | # 2 0.186 | 04:27 | Volume Not Met |
| Breath Volume     | 0.406     | 04:27 |                |
| Air Blank         | 0.000     | 04:27 |                |
| Control Test      | 0.078     | 04:28 |                |
| Air Blank         | 0.000     | 04:28 |                |
| Diagnostics Check | OK        | 04:28 |                |

---

Data Download Date/Time: 05/11/2016 11:49

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MCMINN,KATHY,  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: SIMPKINS,D  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18015080A2  
DRY GAS EXPIRY DATE: 08/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 23:17  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:07

SUBJECT'S LAST NAME: SOLANO

DATE/TIME OF TEST: 04/25/2016 00:33:31

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:27 |
| Air Blank          | 0.000 | 00:27 |
| Control Test       | 0.079 | 00:28 |
| Air Blank          | 0.000 | 00:28 |
| Subject Sample # 1 | 0.125 | 00:29 |
| Breath Volume      | 3.578 | 00:29 |
| Air Blank          | 0.000 | 00:29 |
| Air Blank          | 0.000 | 00:31 |
| Subject Sample # 2 | 0.124 | 00:32 |
| Breath Volume      | 1.777 | 00:32 |
| Air Blank          | 0.000 | 00:32 |
| Control Test       | 0.079 | 00:32 |
| Air Blank          | 0.000 | 00:33 |
| Diagnostics Check  | OK    | 00:33 |

---

Data Download Date/Time: 05/11/2016 11:49

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KIRK,CASEY,S  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: SNAGGS,S  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18015080A2  
DRY GAS EXPIRY DATE: 08/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 00:32  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:57

SUBJECT'S LAST NAME: ACEVEDO VALENCIA

DATE/TIME OF TEST: 04/25/2016 01:24:11

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:18 |
| Air Blank          | 0.000 | 01:18 |
| Control Test       | 0.079 | 01:18 |
| Air Blank          | 0.000 | 01:19 |
| Subject Sample # 1 | 0.137 | 01:19 |
| Breath Volume      | 1.339 | 01:19 |
| Air Blank          | 0.000 | 01:20 |
| Air Blank          | 0.000 | 01:22 |
| Subject Sample # 2 | 0.138 | 01:22 |
| Breath Volume      | 1.410 | 01:22 |
| Air Blank          | 0.000 | 01:23 |
| Control Test       | 0.078 | 01:23 |
| Air Blank          | 0.000 | 01:24 |
| Diagnostics Check  | OK    | 01:24 |

---

Data Download Date/Time: 05/11/2016 11:49

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TORRES,JENIFER,  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: VAN TREESE,S  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18015080A2  
DRY GAS EXPIRY DATE: 08/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 00:43  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:07

SUBJECT'S LAST NAME: TREJO

DATE/TIME OF TEST: 04/27/2016 01:35:33

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:29 |
| Air Blank          | 0.000 | 01:29 |
| Control Test       | 0.079 | 01:29 |
| Air Blank          | 0.000 | 01:30 |
| Subject Sample # 1 | 0.029 | 01:31 |
| Breath Volume      | 4.085 | 01:31 |
| Air Blank          | 0.000 | 01:31 |
| Air Blank          | 0.000 | 01:33 |
| Subject Sample # 2 | 0.029 | 01:34 |
| Breath Volume      | 3.324 | 01:34 |
| Air Blank          | 0.000 | 01:34 |
| Control Test       | 0.079 | 01:34 |
| Air Blank          | 0.000 | 01:35 |
| Diagnostics Check  | OK    | 01:35 |

---

Data Download Date/Time: 05/11/2016 11:49

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BARLAUG, G  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18015080A2  
DRY GAS EXPIRY DATE: 08/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 02:17  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:42

SUBJECT'S LAST NAME: MARGENAT HERNANDEZ

DATE/TIME OF TEST: 04/28/2016 03:09:50

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:02 |
| Air Blank          | 0.000 | 03:03 |
| Control Test       | 0.080 | 03:03 |
| Air Blank          | 0.000 | 03:03 |
| Subject Sample # 1 | 0.161 | 03:04 |
| Breath Volume      | 1.265 | 03:04 |
| Air Blank          | 0.000 | 03:05 |
| Air Blank          | 0.000 | 03:07 |
| Subject Sample # 2 | 0.169 | 03:08 |
| Breath Volume      | 1.722 | 03:08 |
| Air Blank          | 0.000 | 03:08 |
| Control Test       | 0.079 | 03:09 |
| Air Blank          | 0.000 | 03:09 |
| Diagnostics Check  | OK    | 03:09 |

---

Data Download Date/Time: 05/11/2016 11:49

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KIRK,CASEY,S  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: HOBSON,D  
ARRESTING OFFICER AGENCY: U of So FL Campus PD  
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18015080A2  
DRY GAS EXPIRY DATE: 08/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 03:29  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: NORDEN

DATE/TIME OF TEST: 04/29/2016 04:47:27

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:41 |
| Air Blank          | 0.000 | 04:41 |
| Control Test       | 0.080 | 04:42 |
| Air Blank          | 0.000 | 04:42 |
| Subject Sample # 1 | 0.122 | 04:43 |
| Breath Volume      | 2.390 | 04:43 |
| Air Blank          | 0.000 | 04:43 |
| Air Blank          | 0.000 | 04:45 |
| Subject Sample # 2 | 0.121 | 04:45 |
| Breath Volume      | 2.367 | 04:45 |
| Air Blank          | 0.000 | 04:46 |
| Control Test       | 0.079 | 04:46 |
| Air Blank          | 0.000 | 04:47 |
| Diagnostics Check  | OK    | 04:47 |

---

Data Download Date/Time: 05/11/2016 11:49

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MCMINN,KATHY,  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: CAMPBELL,J  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18015080A2  
DRY GAS EXPIRY DATE: 08/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 01:17  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:59

SUBJECT'S LAST NAME: DONLEY

DATE/TIME OF TEST: 04/30/2016 02:27:45

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:19 |
| Air Blank          | 0.000 | 02:19 |
| Control Test       | 0.079 | 02:20 |
| Air Blank          | 0.000 | 02:20 |
| Subject Sample # 1 | 0.185 | 02:21 |
| Breath Volume      | 1.953 | 02:21 |
| Air Blank          | 0.000 | 02:22 |
| Air Blank          | 0.000 | 02:24 |
| Subject Sample # 2 | 0.188 | 02:26 |
| Breath Volume      | 1.320 | 02:26 |
| Air Blank          | 0.000 | 02:26 |
| Control Test       | 0.080 | 02:27 |
| Air Blank          | 0.000 | 02:27 |
| Diagnostics Check  | OK    | 02:27 |

---

Data Download Date/Time: 05/11/2016 11:49



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KIRK,CASEY,S  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: WOODEN,D  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18015080A2  
DRY GAS EXPIRY DATE: 08/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 01:43  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:21

SUBJECT'S LAST NAME: ROJOP SAGASTUME

DATE/TIME OF TEST: 04/30/2016 02:52:08

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:45 |
| Air Blank          | 0.000 | 02:45 |
| Control Test       | 0.078 | 02:45 |
| Air Blank          | 0.000 | 02:46 |
| Subject Sample # 1 | 0.215 | 02:47 |
| Breath Volume      | 2.039 | 02:47 |
| Air Blank          | 0.000 | 02:48 |
| Air Blank          | 0.000 | 02:50 |
| Subject Sample # 2 | 0.210 | 02:50 |
| Breath Volume      | 2.210 | 02:50 |
| Air Blank          | 0.000 | 02:51 |
| Control Test       | 0.078 | 02:51 |
| Air Blank          | 0.000 | 02:52 |
| Diagnostics Check  | OK    | 02:52 |

---

Data Download Date/Time: 05/11/2016 11:50

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CHESSHER,SANDRA,W  
BREATH TEST OPERATOR AGENCY: Lake County SO  
ARRESTING OFFICER: RUBIN,S  
ARRESTING OFFICER AGENCY: Fruitland Park PD  
LAST AGENCY INSPECTION: 03/31/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 12014080A1  
DRY GAS EXPIRY DATE: 06/01/2016  
VIOLATION CODE: DUI  
ARREST TIME: 02:44  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:44

SUBJECT'S LAST NAME: HARTSON JR

DATE/TIME OF TEST: 04/22/2016 04:08:42

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 04:05 |                      |
| Air Blank         | 0.000 | 04:06 |                      |
| Control Test      | 0.078 | 04:06 |                      |
| Air Blank         | 0.000 | 04:06 |                      |
| Subject Sample    | # 1   | 04:07 | Subject Test Refused |
| Breath Volume     | 0.000 | 04:07 |                      |
| Air Blank         | 0.000 | 04:07 |                      |
| Control Test      | 0.078 | 04:08 |                      |
| Air Blank         | 0.000 | 04:08 |                      |
| Diagnostics Check | OK    | 04:08 |                      |

---

Data Download Date/Time: 04/29/2016 22:30

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BOWDEN,MATTHEW,C  
BREATH TEST OPERATOR AGENCY: Lake County SO  
ARRESTING OFFICER: CHESSHER,SANDI  
ARRESTING OFFICER AGENCY: Lake County SO  
LAST AGENCY INSPECTION: 03/31/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 12014080A1  
DRY GAS EXPIRY DATE: 06/01/2016  
VIOLATION CODE: DUI  
ARREST TIME: 21:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: MOHAMMED

DATE/TIME OF TEST: 04/22/2016 21:47:59

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 21:41 |
| Air Blank          | 0.000 | 21:41 |
| Control Test       | 0.077 | 21:41 |
| Air Blank          | 0.000 | 21:42 |
| Subject Sample # 1 | 0.178 | 21:43 |
| Breath Volume      | 4.386 | 21:43 |
| Air Blank          | 0.000 | 21:44 |
| Air Blank          | 0.000 | 21:45 |
| Subject Sample # 2 | 0.175 | 21:46 |
| Breath Volume      | 3.937 | 21:46 |
| Air Blank          | 0.000 | 21:47 |
| Control Test       | 0.077 | 21:47 |
| Air Blank          | 0.000 | 21:47 |
| Diagnostics Check  | OK    | 21:47 |

---

Data Download Date/Time: 04/29/2016 22:30

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CHESSHER,SANDRA,W  
BREATH TEST OPERATOR AGENCY: Lake County SO  
ARRESTING OFFICER: LUKENS,SHAWN  
ARRESTING OFFICER AGENCY: Lake County SO  
LAST AGENCY INSPECTION: 03/31/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 12014080A1  
DRY GAS EXPIRY DATE: 06/01/2016  
VIOLATION CODE: DUI  
ARREST TIME: 23:58  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: MEDEIROS

DATE/TIME OF TEST: 04/24/2016 00:15:15

---

|                   |           |       |            |
|-------------------|-----------|-------|------------|
| Diagnostics Check | OK        | 00:11 |            |
| Air Blank         | 0.000     | 00:11 |            |
| Control Test      | 0.078     | 00:11 |            |
| Air Blank         | 0.000     | 00:12 |            |
| Subject Sample    | # 1 0.130 | 00:12 |            |
| Breath Volume     | 1.218     | 00:12 |            |
| Air Blank         | 0.000     | 00:13 |            |
| Air Blank         |           | 00:14 | RFI Detect |
| Air Blank         |           | 00:15 | RFI Detect |
| Air Blank         |           | 00:15 | RFI Detect |

---

Data Download Date/Time: 04/29/2016 22:30

# **Florida Department of Law Enforcement**

## **Alcohol Testing Program**

### **Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CHESSHER,SANDRA,W  
BREATH TEST OPERATOR AGENCY: Lake County SO  
ARRESTING OFFICER: LUKENS,SHAWN  
ARRESTING OFFICER AGENCY: Lake County SO  
LAST AGENCY INSPECTION: 03/31/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 12014080A1  
DRY GAS EXPIRY DATE: 06/01/2016  
VIOLATION CODE: DUI  
ARREST TIME: 23:58  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: MEDEIROS

DATE/TIME OF TEST: 04/24/2016 00:18:22

---

|                   |       |       |            |
|-------------------|-------|-------|------------|
| Diagnostics Check | OK    | 00:17 |            |
| Air Blank         | 0.000 | 00:17 |            |
| Control Test      | 0.077 | 00:17 |            |
| Air Blank         |       | 00:18 | RFI Detect |
| Air Blank         |       | 00:18 | RFI Detect |

---

Data Download Date/Time: 04/29/2016 22:30

# ***Florida Department of Law Enforcement***

## ***Alcohol Testing Program***

### **Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CHESSHER,SANDRA,W  
BREATH TEST OPERATOR AGENCY: Lake County SO  
ARRESTING OFFICER: LUKENS,SHAWN  
ARRESTING OFFICER AGENCY: Lake County SO  
LAST AGENCY INSPECTION: 03/31/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 12014080A1  
DRY GAS EXPIRY DATE: 06/01/2016  
VIOLATION CODE: DUI  
ARREST TIME: 23:58  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: MEDEIROS

DATE/TIME OF TEST: 04/24/2016 00:22:41

---

|                   |       |       |            |
|-------------------|-------|-------|------------|
| Diagnostics Check | OK    | 00:22 |            |
| Air Blank         |       | 00:22 | RFI Detect |
| Air Blank         | 0.000 | 00:22 |            |

---

Data Download Date/Time: 04/29/2016 22:30

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CHESSHER,SANDRA,W  
BREATH TEST OPERATOR AGENCY: Lake County SO  
ARRESTING OFFICER: LUKENS,SHAWN  
ARRESTING OFFICER AGENCY: Lake County SO  
LAST AGENCY INSPECTION: 03/31/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 12014080A1  
DRY GAS EXPIRY DATE: 06/01/2016  
VIOLATION CODE: DUI  
ARREST TIME: 23:58  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: MEDEIROS

DATE/TIME OF TEST: 04/24/2016 00:30:01

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:23 |
| Air Blank          | 0.000 | 00:24 |
| Control Test       | 0.078 | 00:24 |
| Air Blank          | 0.000 | 00:24 |
| Subject Sample # 1 | 0.133 | 00:25 |
| Breath Volume      | 1.589 | 00:25 |
| Air Blank          | 0.000 | 00:25 |
| Air Blank          | 0.000 | 00:27 |
| Subject Sample # 2 | 0.134 | 00:28 |
| Breath Volume      | 1.835 | 00:28 |
| Air Blank          | 0.000 | 00:29 |
| Control Test       | 0.078 | 00:29 |
| Air Blank          | 0.000 | 00:29 |
| Diagnostics Check  | OK    | 00:30 |

---

Data Download Date/Time: 04/29/2016 22:30

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LUKENS,SHAWN,C  
BREATH TEST OPERATOR AGENCY: Lake County SO  
ARRESTING OFFICER: CHESSHER,SANRA  
ARRESTING OFFICER AGENCY: Lake County SO  
LAST AGENCY INSPECTION: 03/31/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 12014080A1  
DRY GAS EXPIRY DATE: 06/01/2016  
VIOLATION CODE: DUI  
ARREST TIME: 03:45  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: GOMEZ

DATE/TIME OF TEST: 04/24/2016 03:48:35

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | Fail  | 03:48 | Voltage/Current Test |
| Air Blank         | 0.000 | 03:48 |                      |

---

Data Download Date/Time: 04/29/2016 22:30



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LUKENS,SHAWN,C  
BREATH TEST OPERATOR AGENCY: Lake County SO  
ARRESTING OFFICER: CHESSHER,SANDRA  
ARRESTING OFFICER AGENCY: Lake County SO  
LAST AGENCY INSPECTION: 03/31/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 12014080A1  
DRY GAS EXPIRY DATE: 06/01/2016  
VIOLATION CODE: DUI  
ARREST TIME: 03:45  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: GOMEZ

DATE/TIME OF TEST: 04/24/2016 03:50:47

---

|                   |       |       |            |
|-------------------|-------|-------|------------|
| Diagnostics Check | OK    | 03:50 |            |
| Air Blank         |       | 03:50 | RFI Detect |
| Air Blank         | 0.000 | 03:50 |            |

---

Data Download Date/Time: 04/29/2016 22:30

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LUKENS,SHAWN,C  
BREATH TEST OPERATOR AGENCY: Lake County SO  
ARRESTING OFFICER: CHESSHER,SANDRA  
ARRESTING OFFICER AGENCY: Lake County SO  
LAST AGENCY INSPECTION: 03/31/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 12014080A1  
DRY GAS EXPIRY DATE: 06/01/2016  
VIOLATION CODE: DUI  
ARREST TIME: 03:45  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: GOMEZ

DATE/TIME OF TEST: 04/24/2016 03:59:08

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:52 |
| Air Blank          | 0.000 | 03:52 |
| Control Test       | 0.078 | 03:53 |
| Air Blank          | 0.000 | 03:53 |
| Subject Sample # 1 | 0.130 | 03:54 |
| Breath Volume      | 1.648 | 03:54 |
| Air Blank          | 0.000 | 03:54 |
| Air Blank          | 0.000 | 03:56 |
| Subject Sample # 2 | 0.132 | 03:57 |
| Breath Volume      | 1.652 | 03:57 |
| Air Blank          | 0.000 | 03:58 |
| Control Test       | 0.078 | 03:58 |
| Air Blank          | 0.000 | 03:59 |
| Diagnostics Check  | OK    | 03:59 |

---

Data Download Date/Time: 04/29/2016 22:30

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PUSEY,ANTHONY,C  
BREATH TEST OPERATOR AGENCY: Panama City Beach PD  
ARRESTING OFFICER: CHAVIRA,BRANDON  
ARRESTING OFFICER AGENCY: Panama City Beach PD  
LAST AGENCY INSPECTION: 03/28/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1752856  
DRY GAS EXPIRY DATE: 10/13/2017  
VIOLATION CODE: DUI  
ARREST TIME: 01:26  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:38

SUBJECT'S LAST NAME: STEPHENS

DATE/TIME OF TEST: 04/04/2016 02:19:00

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:12 |
| Air Blank          | 0.000 | 02:12 |
| Control Test       | 0.078 | 02:13 |
| Air Blank          | 0.000 | 02:13 |
| Subject Sample # 1 | 0.199 | 02:14 |
| Breath Volume      | 1.113 | 02:14 |
| Air Blank          | 0.000 | 02:15 |
| Air Blank          | 0.000 | 02:16 |
| Subject Sample # 2 | 0.194 | 02:17 |
| Breath Volume      | 1.406 | 02:17 |
| Air Blank          | 0.000 | 02:18 |
| Control Test       | 0.079 | 02:18 |
| Air Blank          | 0.000 | 02:18 |
| Diagnostics Check  | OK    | 02:19 |

---

Data Download Date/Time: 04/28/2016 10:14

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LANCE,CHARLES,R  
BREATH TEST OPERATOR AGENCY: Panama City Beach PD  
ARRESTING OFFICER: LANCE,CHARLES  
ARRESTING OFFICER AGENCY: Panama City Beach PD  
LAST AGENCY INSPECTION: 03/28/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1752856  
DRY GAS EXPIRY DATE: 10/13/2017  
VIOLATION CODE: DUI  
ARREST TIME: 18:46  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:11

SUBJECT'S LAST NAME: FONDREN

DATE/TIME OF TEST: 04/08/2016 19:38:43

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 19:31 |
| Air Blank          | 0.000 | 19:31 |
| Control Test       | 0.080 | 19:32 |
| Air Blank          | 0.000 | 19:32 |
| Subject Sample # 1 | 0.217 | 19:33 |
| Breath Volume      | 2.031 | 19:33 |
| Air Blank          | 0.000 | 19:34 |
| Air Blank          | 0.000 | 19:36 |
| Subject Sample # 2 | 0.205 | 19:37 |
| Breath Volume      | 1.371 | 19:37 |
| Air Blank          | 0.000 | 19:37 |
| Control Test       | 0.079 | 19:38 |
| Air Blank          | 0.000 | 19:38 |
| Diagnostics Check  | OK    | 19:38 |

---

Data Download Date/Time: 04/28/2016 10:14

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PUSEY,ANTHONY,C  
BREATH TEST OPERATOR AGENCY: Panama City Beach PD  
ARRESTING OFFICER: PUSEY,ANTHONY  
ARRESTING OFFICER AGENCY: Panama City Beach PD  
LAST AGENCY INSPECTION: 03/28/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1752856  
DRY GAS EXPIRY DATE: 10/13/2017  
VIOLATION CODE: DUI  
ARREST TIME: 05:09  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:37

SUBJECT'S LAST NAME: HIGGS

DATE/TIME OF TEST: 04/09/2016 06:27:53

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 06:20 |
| Air Blank          | 0.000 | 06:21 |
| Control Test       | 0.081 | 06:21 |
| Air Blank          | 0.000 | 06:21 |
| Subject Sample # 1 | 0.143 | 06:23 |
| Breath Volume      | 2.351 | 06:23 |
| Air Blank          | 0.000 | 06:23 |
| Air Blank          | 0.000 | 06:25 |
| Subject Sample # 2 | 0.144 | 06:26 |
| Breath Volume      | 1.984 | 06:26 |
| Air Blank          | 0.000 | 06:26 |
| Control Test       | 0.080 | 06:27 |
| Air Blank          | 0.000 | 06:27 |
| Diagnostics Check  | OK    | 06:27 |

---

Data Download Date/Time: 04/28/2016 10:14

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PUSEY,ANTHONY,C  
BREATH TEST OPERATOR AGENCY: Panama City Beach PD  
ARRESTING OFFICER: HOWARD,NOEL  
ARRESTING OFFICER AGENCY: Panama City Beach PD  
LAST AGENCY INSPECTION: 03/28/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1752856  
DRY GAS EXPIRY DATE: 10/13/2017  
VIOLATION CODE: DUI  
ARREST TIME: 21:34  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: LEAVINS

DATE/TIME OF TEST: 04/09/2016 22:14:56

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:08 |
| Air Blank          | 0.000 | 22:09 |
| Control Test       | 0.081 | 22:09 |
| Air Blank          | 0.000 | 22:10 |
| Subject Sample # 1 | 0.000 | 22:10 |
| Breath Volume      | 1.253 | 22:10 |
| Air Blank          | 0.000 | 22:11 |
| Air Blank          | 0.000 | 22:13 |
| Subject Sample # 2 | 0.000 | 22:13 |
| Breath Volume      | 1.472 | 22:13 |
| Air Blank          | 0.000 | 22:14 |
| Control Test       | 0.080 | 22:14 |
| Air Blank          | 0.000 | 22:14 |
| Diagnostics Check  | OK    | 22:14 |

---

Data Download Date/Time: 04/28/2016 10:14

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PUSEY,ANTHONY,C  
BREATH TEST OPERATOR AGENCY: Panama City Beach PD  
ARRESTING OFFICER: MOSS,ROBERT  
ARRESTING OFFICER AGENCY: Panama City Beach PD  
LAST AGENCY INSPECTION: 03/28/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1752856  
DRY GAS EXPIRY DATE: 10/13/2017  
VIOLATION CODE: DUI  
ARREST TIME: 00:29  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:29

SUBJECT'S LAST NAME: KAMM

DATE/TIME OF TEST: 04/14/2016 00:57:16

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:50 |
| Air Blank          | 0.000 | 00:50 |
| Control Test       | 0.080 | 00:50 |
| Air Blank          | 0.000 | 00:51 |
| Subject Sample # 1 | 0.194 | 00:52 |
| Breath Volume      | 2.035 | 00:52 |
| Air Blank          | 0.000 | 00:52 |
| Air Blank          | 0.000 | 00:54 |
| Subject Sample # 2 | 0.195 | 00:55 |
| Breath Volume      | 2.054 | 00:55 |
| Air Blank          | 0.000 | 00:56 |
| Control Test       | 0.079 | 00:56 |
| Air Blank          | 0.000 | 00:57 |
| Diagnostics Check  | OK    | 00:57 |

---

Data Download Date/Time: 04/28/2016 10:14

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LANCE,CHARLES,R  
BREATH TEST OPERATOR AGENCY: Panama City Beach PD  
ARRESTING OFFICER: LANCE,CHARLES  
ARRESTING OFFICER AGENCY: Panama City Beach PD  
LAST AGENCY INSPECTION: 03/28/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1752856  
DRY GAS EXPIRY DATE: 10/13/2017  
VIOLATION CODE: DUI  
ARREST TIME: 16:38  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:44

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 04/14/2016 17:11:03

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 17:04 |
| Air Blank          | 0.000 | 17:05 |
| Control Test       | 0.081 | 17:05 |
| Air Blank          | 0.000 | 17:06 |
| Subject Sample # 1 | 0.308 | 17:06 |
| Breath Volume      | 2.929 | 17:06 |
| Air Blank          | 0.000 | 17:07 |
| Air Blank          | 0.000 | 17:08 |
| Subject Sample # 2 | 0.314 | 17:09 |
| Breath Volume      | 1.992 | 17:09 |
| Air Blank          | 0.000 | 17:10 |
| Control Test       | 0.079 | 17:10 |
| Air Blank          | 0.000 | 17:10 |
| Diagnostics Check  | OK    | 17:11 |

---

Data Download Date/Time: 04/28/2016 10:14



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JONES,CASEY,L  
BREATH TEST OPERATOR AGENCY: Panama City Beach PD  
ARRESTING OFFICER: MOSS,ROBERT  
ARRESTING OFFICER AGENCY: Panama City Beach PD  
LAST AGENCY INSPECTION: 03/28/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1752856  
DRY GAS EXPIRY DATE: 10/13/2017  
VIOLATION CODE: DUI  
ARREST TIME: 00:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: MULFORD

DATE/TIME OF TEST: 04/16/2016 01:34:06

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:26 |
| Air Blank          | 0.000 | 01:27 |
| Control Test       | 0.080 | 01:27 |
| Air Blank          | 0.000 | 01:27 |
| Subject Sample # 1 | 0.141 | 01:29 |
| Breath Volume      | 2.433 | 01:29 |
| Air Blank          | 0.000 | 01:30 |
| Air Blank          | 0.000 | 01:32 |
| Subject Sample # 2 | 0.139 | 01:32 |
| Breath Volume      | 1.324 | 01:32 |
| Air Blank          | 0.000 | 01:33 |
| Control Test       | 0.079 | 01:33 |
| Air Blank          | 0.000 | 01:34 |
| Diagnostics Check  | OK    | 01:34 |

---

Data Download Date/Time: 04/28/2016 10:14

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PELFREY,PAUL,D  
BREATH TEST OPERATOR AGENCY: Panama City Beach PD  
ARRESTING OFFICER: PELFREY,PAUL  
ARRESTING OFFICER AGENCY: Panama City Beach PD  
LAST AGENCY INSPECTION: 03/28/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1752856  
DRY GAS EXPIRY DATE: 10/13/2017  
VIOLATION CODE: DUI  
ARREST TIME: 00:00  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: SPEED

DATE/TIME OF TEST: 04/17/2016 00:37:52

---

|                   |       |       |            |
|-------------------|-------|-------|------------|
| Diagnostics Check | OK    | 00:37 |            |
| Air Blank         |       | 00:37 | RFI Detect |
| Air Blank         | 0.000 | 00:37 |            |

---

Data Download Date/Time: 04/28/2016 10:14

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PELFREY,PAUL,D  
BREATH TEST OPERATOR AGENCY: Panama City Beach PD  
ARRESTING OFFICER: PELFREY,PAUL  
ARRESTING OFFICER AGENCY: Panama City Beach PD  
LAST AGENCY INSPECTION: 03/28/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1752856  
DRY GAS EXPIRY DATE: 10/13/2017  
VIOLATION CODE: DUI  
ARREST TIME: 00:00  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: SPEED

DATE/TIME OF TEST: 04/17/2016 00:53:01

---

|                   |           |       |                    |
|-------------------|-----------|-------|--------------------|
| Diagnostics Check | OK        | 00:41 |                    |
| Air Blank         | 0.000     | 00:42 |                    |
| Control Test      | 0.078     | 00:42 |                    |
| Air Blank         | 0.000     | 00:42 |                    |
| Subject Sample    | # 1 0.197 | 00:45 | Volume Not Met     |
| Breath Volume     | 0.019     | 00:45 |                    |
| Air Blank         | 0.000     | 00:46 |                    |
| Air Blank         | 0.000     | 00:48 |                    |
| Subject Sample    | # 2       | 00:51 | No Sample Provided |
| Breath Volume     | 0.000     | 00:51 |                    |
| Air Blank         | 0.000     | 00:52 |                    |
| Control Test      | 0.078     | 00:52 |                    |
| Air Blank         | 0.000     | 00:52 |                    |
| Diagnostics Check | OK        | 00:53 |                    |

---

Data Download Date/Time: 04/28/2016 10:14

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LANCE,CHARLES,R  
BREATH TEST OPERATOR AGENCY: Panama City Beach PD  
ARRESTING OFFICER: LANCE,CHARLES  
ARRESTING OFFICER AGENCY: Panama City Beach PD  
LAST AGENCY INSPECTION: 03/28/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1752856  
DRY GAS EXPIRY DATE: 10/13/2017  
VIOLATION CODE: DUI  
ARREST TIME: 17:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:25

SUBJECT'S LAST NAME: ALTENBURG

DATE/TIME OF TEST: 04/24/2016 17:53:54

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 17:46 |
| Air Blank          | 0.000 | 17:46 |
| Control Test       | 0.079 | 17:47 |
| Air Blank          | 0.000 | 17:47 |
| Subject Sample # 1 | 0.257 | 17:48 |
| Breath Volume      | 2.210 | 17:48 |
| Air Blank          | 0.000 | 17:48 |
| Air Blank          | 0.000 | 17:50 |
| Subject Sample # 2 | 0.257 | 17:52 |
| Breath Volume      | 1.695 | 17:52 |
| Air Blank          | 0.000 | 17:53 |
| Control Test       | 0.079 | 17:53 |
| Air Blank          | 0.000 | 17:53 |
| Diagnostics Check  | OK    | 17:53 |

---

Data Download Date/Time: 04/28/2016 10:14

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CANNON,JAMES,J  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: HAMILTON-REYES,CORTNEY  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 03/09/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03415080A2  
DRY GAS EXPIRY DATE: 03/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 18:51  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:14

SUBJECT'S LAST NAME: WAKEMAN

DATE/TIME OF TEST: 04/01/2016 19:43:14

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 19:35 |
| Air Blank          | 0.000 | 19:36 |
| Control Test       | 0.081 | 19:36 |
| Air Blank          | 0.000 | 19:37 |
| Subject Sample # 1 | 0.091 | 19:37 |
| Breath Volume      | 1.171 | 19:37 |
| Air Blank          | 0.000 | 19:38 |
| Air Blank          | 0.000 | 19:40 |
| Subject Sample # 2 | 0.087 | 19:41 |
| Breath Volume      | 1.589 | 19:41 |
| Air Blank          | 0.000 | 19:42 |
| Control Test       | 0.081 | 19:42 |
| Air Blank          | 0.000 | 19:43 |
| Diagnostics Check  | OK    | 19:43 |

---

Data Download Date/Time: 04/13/2016 08:52

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CANNON,JAMES,J  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: HAMILTON-REYES,CORTNEY  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 03/09/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03415080A2  
DRY GAS EXPIRY DATE: 03/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 01:27  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:44

SUBJECT'S LAST NAME: MONTES-LAWRENCE

DATE/TIME OF TEST: 04/07/2016 02:11:26

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:05 |
| Air Blank          | 0.000 | 02:05 |
| Control Test       | 0.080 | 02:06 |
| Air Blank          | 0.000 | 02:06 |
| Subject Sample # 1 | 0.000 | 02:07 |
| Breath Volume      | 2.410 | 02:07 |
| Air Blank          | 0.000 | 02:07 |
| Air Blank          | 0.000 | 02:09 |
| Subject Sample # 2 | 0.000 | 02:10 |
| Breath Volume      | 1.722 | 02:10 |
| Air Blank          | 0.000 | 02:10 |
| Control Test       | 0.080 | 02:10 |
| Air Blank          | 0.000 | 02:11 |
| Diagnostics Check  | OK    | 02:11 |

---

Data Download Date/Time: 04/13/2016 08:52

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CANNON,JAMES,J  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: ROSAS,JASON  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 03/09/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03415080A2  
DRY GAS EXPIRY DATE: 03/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 02:53  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: DORR-RUOT

DATE/TIME OF TEST: 04/08/2016 03:47:16

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:40 |
| Air Blank          | 0.000 | 03:40 |
| Control Test       | 0.080 | 03:41 |
| Air Blank          | 0.000 | 03:41 |
| Subject Sample # 1 | 0.148 | 03:42 |
| Breath Volume      | 1.621 | 03:42 |
| Air Blank          | 0.000 | 03:42 |
| Air Blank          | 0.000 | 03:44 |
| Subject Sample # 2 | 0.142 | 03:45 |
| Breath Volume      | 1.355 | 03:45 |
| Air Blank          | 0.000 | 03:46 |
| Control Test       | 0.081 | 03:46 |
| Air Blank          | 0.000 | 03:47 |
| Diagnostics Check  | OK    | 03:47 |

---

Data Download Date/Time: 04/13/2016 08:52

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ESTERLE,HUGH,L  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: SCARLATO,FRANK  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 04/13/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03415080A2  
DRY GAS EXPIRY DATE: 03/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 19:04  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:18

SUBJECT'S LAST NAME: DUFF

DATE/TIME OF TEST: 04/15/2016 19:44:53

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 19:38 |
| Air Blank          | 0.000 | 19:39 |
| Control Test       | 0.081 | 19:39 |
| Air Blank          | 0.000 | 19:39 |
| Subject Sample # 1 | 0.206 | 19:40 |
| Breath Volume      | 2.390 | 19:40 |
| Air Blank          | 0.000 | 19:40 |
| Air Blank          | 0.000 | 19:42 |
| Subject Sample # 2 | 0.208 | 19:43 |
| Breath Volume      | 3.066 | 19:43 |
| Air Blank          | 0.000 | 19:43 |
| Control Test       | 0.080 | 19:44 |
| Air Blank          | 0.000 | 19:44 |
| Diagnostics Check  | OK    | 19:44 |

---

Data Download Date/Time: 05/17/2016 10:50



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CAUSER,SGT,H  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: CRAGIN,S  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 04/13/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03415080A2  
DRY GAS EXPIRY DATE: 03/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 00:07  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: STAHL

DATE/TIME OF TEST: 04/16/2016 00:53:14

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:46 |
| Air Blank          | 0.000 | 00:46 |
| Control Test       | 0.081 | 00:46 |
| Air Blank          | 0.000 | 00:47 |
| Subject Sample # 1 | 0.178 | 00:48 |
| Breath Volume      | 1.996 | 00:48 |
| Air Blank          | 0.000 | 00:49 |
| Air Blank          | 0.000 | 00:50 |
| Subject Sample # 2 | 0.169 | 00:51 |
| Breath Volume      | 1.613 | 00:51 |
| Air Blank          | 0.000 | 00:52 |
| Control Test       | 0.080 | 00:52 |
| Air Blank          | 0.000 | 00:53 |
| Diagnostics Check  | OK    | 00:53 |

---

Data Download Date/Time: 05/17/2016 10:50

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FESSEL,DANIEL,C  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: KORTRIGHT,THOMAS  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 04/13/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03415080A2  
DRY GAS EXPIRY DATE: 03/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 01:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:31

SUBJECT'S LAST NAME: IVERSON

DATE/TIME OF TEST: 04/17/2016 02:03:41

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:56 |
| Air Blank          | 0.000 | 01:57 |
| Control Test       | 0.081 | 01:57 |
| Air Blank          | 0.000 | 01:58 |
| Subject Sample # 1 | 0.126 | 01:59 |
| Breath Volume      | 1.355 | 01:59 |
| Air Blank          | 0.000 | 01:59 |
| Air Blank          | 0.000 | 02:01 |
| Subject Sample # 2 | 0.137 | 02:02 |
| Breath Volume      | 1.574 | 02:02 |
| Air Blank          | 0.000 | 02:02 |
| Control Test       | 0.080 | 02:03 |
| Air Blank          | 0.000 | 02:03 |
| Diagnostics Check  | OK    | 02:03 |

---

Data Download Date/Time: 05/17/2016 10:50

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CANNON,JAMES,J  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: TOBLER,AMANDA  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 04/13/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03415080A2  
DRY GAS EXPIRY DATE: 03/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 04:39  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:49

SUBJECT'S LAST NAME: QUINONES

DATE/TIME OF TEST: 04/17/2016 05:30:20

---

|                   |           |       |                |
|-------------------|-----------|-------|----------------|
| Diagnostics Check | OK        | 05:15 |                |
| Air Blank         | 0.000     | 05:15 |                |
| Control Test      | 0.080     | 05:15 |                |
| Air Blank         | 0.000     | 05:16 |                |
| Subject Sample    | # 1 0.129 | 05:19 |                |
| Breath Volume     | 1.480     | 05:19 |                |
| Air Blank         | 0.000     | 05:19 |                |
| Air Blank         | 0.000     | 05:21 |                |
| Subject Sample    | # 2 0.095 | 05:24 | Volume Not Met |
| Breath Volume     | 0.160     | 05:24 |                |
| Air Blank         | 0.000     | 05:25 |                |
| Air Blank         | 0.000     | 05:27 |                |
| Subject Sample    | # 3 0.136 | 05:28 |                |
| Breath Volume     | 1.386     | 05:28 |                |
| Air Blank         | 0.000     | 05:29 |                |
| Control Test      | 0.080     | 05:29 |                |
| Air Blank         | 0.000     | 05:30 |                |
| Diagnostics Check | OK        | 05:30 |                |

---

Data Download Date/Time: 05/17/2016 10:50

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CAUSER,SGT,H  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: JOHNSON,S  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 04/13/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03415080A2  
DRY GAS EXPIRY DATE: 03/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 03:14  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:42

SUBJECT'S LAST NAME: FUHRMAN

DATE/TIME OF TEST: 04/18/2016 04:10:36

---

|                   |           |       |            |
|-------------------|-----------|-------|------------|
| Diagnostics Check | OK        | 04:04 |            |
| Air Blank         | 0.000     | 04:04 |            |
| Control Test      | 0.080     | 04:05 |            |
| Air Blank         | 0.000     | 04:05 |            |
| Subject Sample    | # 1 0.192 | 04:06 |            |
| Breath Volume     | 2.765     | 04:06 |            |
| Air Blank         | 0.000     | 04:06 |            |
| Air Blank         | 0.000     | 04:08 |            |
| Subject Sample    | # 2 0.180 | 04:09 |            |
| Breath Volume     | 1.917     | 04:09 |            |
| Air Blank         | 0.000     | 04:10 |            |
| Control Test      | 0.080     | 04:10 |            |
| Air Blank         |           | 04:10 | RFI Detect |
| Air Blank         |           | 04:10 | RFI Detect |

---

Data Download Date/Time: 05/17/2016 10:50

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CAUSER,SGT,H  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: JOHNSON,S  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 04/13/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03415080A2  
DRY GAS EXPIRY DATE: 03/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 03:14  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:42

SUBJECT'S LAST NAME: FUHRMAN

DATE/TIME OF TEST: 04/18/2016 04:20:09

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:13 |
| Air Blank          | 0.000 | 04:13 |
| Control Test       | 0.080 | 04:13 |
| Air Blank          | 0.000 | 04:14 |
| Subject Sample # 1 | 0.179 | 04:15 |
| Breath Volume      | 2.894 | 04:15 |
| Air Blank          | 0.000 | 04:16 |
| Air Blank          | 0.000 | 04:18 |
| Subject Sample # 2 | 0.179 | 04:18 |
| Breath Volume      | 3.007 | 04:18 |
| Air Blank          | 0.000 | 04:19 |
| Control Test       | 0.080 | 04:19 |
| Air Blank          | 0.000 | 04:20 |
| Diagnostics Check  | OK    | 04:20 |

---

Data Download Date/Time: 05/17/2016 10:50

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARKLEY,JEFFREY,G  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: FILLAK,MICHAEL  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 04/13/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03415080A2  
DRY GAS EXPIRY DATE: 03/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 01:07  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:27

SUBJECT'S LAST NAME: DORANGE

DATE/TIME OF TEST: 04/19/2016 01:52:36

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 01:47 |                      |
| Air Blank         | 0.000 | 01:48 |                      |
| Control Test      | 0.080 | 01:48 |                      |
| Air Blank         | 0.000 | 01:48 |                      |
| Subject Sample    | # 1   | 01:51 | Subject Test Refused |
| Breath Volume     | 0.000 | 01:51 |                      |
| Air Blank         | 0.000 | 01:51 |                      |
| Control Test      | 0.081 | 01:52 |                      |
| Air Blank         | 0.000 | 01:52 |                      |
| Diagnostics Check | OK    | 01:52 |                      |

---

Data Download Date/Time: 05/17/2016 10:50

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CAUSER,SGT,H  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: ALEY,K  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 04/13/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03415080A2  
DRY GAS EXPIRY DATE: 03/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 03:48  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:01

SUBJECT'S LAST NAME: FORMOSA

DATE/TIME OF TEST: 04/22/2016 04:30:27

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:23 |
| Air Blank          | 0.000 | 04:24 |
| Control Test       | 0.080 | 04:24 |
| Air Blank          | 0.000 | 04:25 |
| Subject Sample # 1 | 0.160 | 04:25 |
| Breath Volume      | 3.019 | 04:25 |
| Air Blank          | 0.000 | 04:26 |
| Air Blank          | 0.000 | 04:28 |
| Subject Sample # 2 | 0.156 | 04:28 |
| Breath Volume      | 3.234 | 04:28 |
| Air Blank          | 0.000 | 04:29 |
| Control Test       | 0.080 | 04:29 |
| Air Blank          | 0.000 | 04:30 |
| Diagnostics Check  | OK    | 04:30 |

---

Data Download Date/Time: 05/17/2016 10:50

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HEARN,JUSTIN,M  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: VENSKI,JASON  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 04/13/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03415080A2  
DRY GAS EXPIRY DATE: 03/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 01:48  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:59

SUBJECT'S LAST NAME: CHARLES THOMPSON

DATE/TIME OF TEST: 04/24/2016 02:31:00

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:24 |
| Air Blank          | 0.000 | 02:25 |
| Control Test       | 0.081 | 02:25 |
| Air Blank          | 0.000 | 02:26 |
| Subject Sample # 1 | 0.071 | 02:26 |
| Breath Volume      | 2.835 | 02:26 |
| Air Blank          | 0.000 | 02:27 |
| Air Blank          | 0.000 | 02:29 |
| Subject Sample # 2 | 0.072 | 02:29 |
| Breath Volume      | 3.753 | 02:29 |
| Air Blank          | 0.000 | 02:30 |
| Control Test       | 0.080 | 02:30 |
| Air Blank          | 0.000 | 02:30 |
| Diagnostics Check  | OK    | 02:31 |

---

Data Download Date/Time: 05/17/2016 10:50



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ZARRILLO,N,  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: HEARN,J  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 04/13/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03415080A2  
DRY GAS EXPIRY DATE: 03/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 20:15  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 20:37

SUBJECT'S LAST NAME: FISHER

DATE/TIME OF TEST: 04/27/2016 21:03:31

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 20:57 |
| Air Blank          | 0.000 | 20:57 |
| Control Test       | 0.081 | 20:57 |
| Air Blank          | 0.000 | 20:58 |
| Subject Sample # 1 | 0.212 | 20:59 |
| Breath Volume      | 2.519 | 20:59 |
| Air Blank          | 0.000 | 20:59 |
| Air Blank          | 0.000 | 21:01 |
| Subject Sample # 2 | 0.209 | 21:02 |
| Breath Volume      | 1.996 | 21:02 |
| Air Blank          | 0.000 | 21:02 |
| Control Test       | 0.081 | 21:02 |
| Air Blank          | 0.000 | 21:03 |
| Diagnostics Check  | OK    | 21:03 |

---

Data Download Date/Time: 05/17/2016 10:50

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FILLAK, MICHAEL, O  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: HEARN, JUSTIN  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 04/13/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03415080A2  
DRY GAS EXPIRY DATE: 03/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 02:37  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:47

SUBJECT'S LAST NAME: DEWITT

DATE/TIME OF TEST: 04/29/2016 03:14:42

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:07 |
| Air Blank          | 0.000 | 03:08 |
| Control Test       | 0.081 | 03:08 |
| Air Blank          | 0.000 | 03:09 |
| Subject Sample # 1 | 0.182 | 03:10 |
| Breath Volume      | 3.355 | 03:10 |
| Air Blank          | 0.000 | 03:10 |
| Air Blank          | 0.000 | 03:12 |
| Subject Sample # 2 | 0.177 | 03:13 |
| Breath Volume      | 3.570 | 03:13 |
| Air Blank          | 0.000 | 03:13 |
| Control Test       | 0.081 | 03:14 |
| Air Blank          | 0.000 | 03:14 |
| Diagnostics Check  | OK    | 03:14 |

---

Data Download Date/Time: 05/17/2016 10:51

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FILLAK, MICHAEL, O  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: WILSON, PERRY  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 04/13/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03415080A2  
DRY GAS EXPIRY DATE: 03/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 02:37  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:06

SUBJECT'S LAST NAME: COHEN

DATE/TIME OF TEST: 04/29/2016 03:35:22

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:28 |
| Air Blank          | 0.000 | 03:28 |
| Control Test       | 0.080 | 03:29 |
| Air Blank          | 0.000 | 03:29 |
| Subject Sample # 1 | 0.132 | 03:30 |
| Breath Volume      | 4.050 | 03:30 |
| Air Blank          | 0.000 | 03:31 |
| Air Blank          | 0.000 | 03:32 |
| Subject Sample # 2 | 0.136 | 03:33 |
| Breath Volume      | 2.957 | 03:33 |
| Air Blank          | 0.000 | 03:34 |
| Control Test       | 0.081 | 03:34 |
| Air Blank          | 0.000 | 03:35 |
| Diagnostics Check  | OK    | 03:35 |

---

Data Download Date/Time: 05/17/2016 10:51

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KOLAK,ALLAN,S  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: DURHAM,KARI  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 04/13/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03415080A2  
DRY GAS EXPIRY DATE: 03/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 20:39  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:03

SUBJECT'S LAST NAME: MIDYETT

DATE/TIME OF TEST: 04/29/2016 21:29:38

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 21:23 |
| Air Blank          | 0.000 | 21:23 |
| Control Test       | 0.080 | 21:23 |
| Air Blank          | 0.000 | 21:24 |
| Subject Sample # 1 | 0.193 | 21:25 |
| Breath Volume      | 2.992 | 21:25 |
| Air Blank          | 0.000 | 21:25 |
| Air Blank          | 0.000 | 21:27 |
| Subject Sample # 2 | 0.202 | 21:28 |
| Breath Volume      | 2.871 | 21:28 |
| Air Blank          | 0.000 | 21:28 |
| Control Test       | 0.080 | 21:29 |
| Air Blank          | 0.000 | 21:29 |
| Diagnostics Check  | OK    | 21:29 |

---

Data Download Date/Time: 05/17/2016 10:51

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SUSKEVICH,CURT,E  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: THOMAS,D  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 04/13/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03415080A2  
DRY GAS EXPIRY DATE: 03/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 23:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:16

SUBJECT'S LAST NAME: PEREZ PADILLA

DATE/TIME OF TEST: 04/29/2016 23:50:30

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:44 |
| Air Blank          | 0.000 | 23:44 |
| Control Test       | 0.080 | 23:44 |
| Air Blank          | 0.000 | 23:45 |
| Subject Sample # 1 | 0.205 | 23:45 |
| Breath Volume      | 2.082 | 23:45 |
| Air Blank          | 0.000 | 23:46 |
| Air Blank          | 0.000 | 23:48 |
| Subject Sample # 2 | 0.205 | 23:49 |
| Breath Volume      | 2.378 | 23:49 |
| Air Blank          | 0.000 | 23:49 |
| Control Test       | 0.081 | 23:49 |
| Air Blank          | 0.000 | 23:50 |
| Diagnostics Check  | OK    | 23:50 |

---

Data Download Date/Time: 05/17/2016 10:51

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FLINT,SGT,C  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: FLINT,SGT  
ARRESTING OFFICER AGENCY: Sarasota County SO  
LAST AGENCY INSPECTION: 03/25/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28814080A3  
DRY GAS EXPIRY DATE: 11/05/2016  
VIOLATION CODE: DUI  
ARREST TIME: 23:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:02

SUBJECT'S LAST NAME: BOHANNON

DATE/TIME OF TEST: 04/23/2016 23:32:09

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:23 |
| Air Blank          | 0.000 | 23:23 |
| Control Test       | 0.080 | 23:23 |
| Air Blank          | 0.000 | 23:24 |
| Subject Sample # 1 | 0.220 | 23:27 |
| Breath Volume      | 1.644 | 23:27 |
| Air Blank          | 0.000 | 23:28 |
| Air Blank          | 0.000 | 23:29 |
| Subject Sample # 2 | 0.237 | 23:30 |
| Breath Volume      | 1.226 | 23:30 |
| Air Blank          | 0.000 | 23:31 |
| Control Test       | 0.078 | 23:31 |
| Air Blank          | 0.000 | 23:32 |
| Diagnostics Check  | OK    | 23:32 |

---

Data Download Date/Time: 04/28/2016 10:27

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MINGS,DEPUTY G.,  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: MINGS,DEPUTY G.  
ARRESTING OFFICER AGENCY: Sarasota County SO  
LAST AGENCY INSPECTION: 03/25/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28814080A3  
DRY GAS EXPIRY DATE: 11/05/2016  
VIOLATION CODE: DUI  
ARREST TIME: 02:57  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 04/24/2016 03:12:15

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:05 |
| Air Blank          | 0.000 | 03:06 |
| Control Test       | 0.080 | 03:06 |
| Air Blank          | 0.000 | 03:07 |
| Subject Sample # 1 | 0.190 | 03:07 |
| Breath Volume      | 2.167 | 03:07 |
| Air Blank          | 0.000 | 03:08 |
| Air Blank          | 0.000 | 03:10 |
| Subject Sample # 2 | 0.182 | 03:10 |
| Breath Volume      | 2.132 | 03:10 |
| Air Blank          | 0.000 | 03:11 |
| Control Test       | 0.079 | 03:11 |
| Air Blank          | 0.000 | 03:12 |
| Diagnostics Check  | OK    | 03:12 |

---

Data Download Date/Time: 04/28/2016 10:27

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BRADY,JOSHUA,J  
BREATH TEST OPERATOR AGENCY: Monroe County SO  
ARRESTING OFFICER: HEIN,PAUL  
ARRESTING OFFICER AGENCY: FFWCC (West Palm)  
LAST AGENCY INSPECTION: 03/10/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1816818  
DRY GAS EXPIRY DATE: 02/13/2018  
VIOLATION CODE: BUI  
ARREST TIME: 18:51  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 19:38

SUBJECT'S LAST NAME: VALDEZ

DATE/TIME OF TEST: 04/02/2016 20:02:06

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 19:59 |                      |
| Air Blank         | 0.000 | 19:59 |                      |
| Control Test      | 0.081 | 19:59 |                      |
| Air Blank         | 0.000 | 20:00 |                      |
| Subject Sample    | # 1   | 20:00 | Subject Test Refused |
| Breath Volume     | 0.000 | 20:00 |                      |
| Air Blank         | 0.000 | 20:01 |                      |
| Control Test      | 0.081 | 20:01 |                      |
| Air Blank         | 0.000 | 20:02 |                      |
| Diagnostics Check | OK    | 20:02 |                      |

---

Data Download Date/Time: 04/28/2016 09:19



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TORRES,DEANNA,N  
BREATH TEST OPERATOR AGENCY: Monroe County SO  
ARRESTING OFFICER: MADNICK,JASON  
ARRESTING OFFICER AGENCY: Monroe County SO  
LAST AGENCY INSPECTION: 03/10/2016

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 1816818  
DRY GAS EXPIRY DATE: 02/13/2018  
VIOLATION CODE: DUI  
ARREST TIME: 02:09  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:38

SUBJECT'S LAST NAME: KAIGLER

DATE/TIME OF TEST: 04/06/2016 03:04:55

---

|                   |           |       |                 |
|-------------------|-----------|-------|-----------------|
| Diagnostics Check | OK        | 02:58 |                 |
| Air Blank         | 0.000     | 02:59 |                 |
| Control Test      | 0.081     | 02:59 |                 |
| Air Blank         | 0.000     | 03:00 |                 |
| Subject Sample    | # 1 0.129 | 03:00 |                 |
| Breath Volume     | 1.230     | 03:00 |                 |
| Air Blank         | 0.000     | 03:01 |                 |
| Air Blank         | 0.000     | 03:03 |                 |
| Subject Sample    | # 2 0.116 | 03:04 |                 |
| Breath Volume     | 2.117     | 03:04 |                 |
| Air Blank         |           | 03:04 | Improper Sample |
| Air Blank         | 0.000     | 03:04 |                 |

---

Data Download Date/Time: 04/28/2016 09:19

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KEITH,JASON,T  
BREATH TEST OPERATOR AGENCY: Monroe County SO  
ARRESTING OFFICER: SANCHEZ,LUIS  
ARRESTING OFFICER AGENCY: Monroe County SO  
LAST AGENCY INSPECTION: 03/10/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1816818  
DRY GAS EXPIRY DATE: 02/13/2018  
VIOLATION CODE: DUI  
ARREST TIME: 23:42  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:08

SUBJECT'S LAST NAME: ROUSE

DATE/TIME OF TEST: 04/09/2016 00:34:36

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 00:31 |                      |
| Air Blank         | 0.000 | 00:31 |                      |
| Control Test      | 0.080 | 00:31 |                      |
| Air Blank         | 0.000 | 00:32 |                      |
| Subject Sample    | # 1   | 00:33 | Subject Test Refused |
| Breath Volume     | 0.000 | 00:33 |                      |
| Air Blank         | 0.000 | 00:33 |                      |
| Control Test      | 0.081 | 00:34 |                      |
| Air Blank         | 0.000 | 00:34 |                      |
| Diagnostics Check | OK    | 00:34 |                      |

---

Data Download Date/Time: 04/28/2016 09:19

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HANEY,MICHAEL,P  
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)  
ARRESTING OFFICER: HEIN,PAUL  
ARRESTING OFFICER AGENCY: FFWCC (West Palm)  
LAST AGENCY INSPECTION: 03/10/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1816818  
DRY GAS EXPIRY DATE: 02/13/2018  
VIOLATION CODE: BUI  
ARREST TIME: 16:27  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:35

SUBJECT'S LAST NAME: DISSELKOEN

DATE/TIME OF TEST: 04/09/2016 19:02:31

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 18:57 |                      |
| Air Blank         | 0.000 | 18:57 |                      |
| Control Test      | 0.080 | 18:58 |                      |
| Air Blank         | 0.000 | 18:58 |                      |
| Subject Sample    | # 1   | 19:01 | Subject Test Refused |
| Breath Volume     | 0.000 | 19:01 |                      |
| Air Blank         | 0.000 | 19:01 |                      |
| Control Test      | 0.081 | 19:01 |                      |
| Air Blank         | 0.000 | 19:02 |                      |
| Diagnostics Check | OK    | 19:02 |                      |

---

Data Download Date/Time: 04/28/2016 09:19

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HANEY,MICHAEL,P  
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)  
ARRESTING OFFICER: HANEY,MICHAEL  
ARRESTING OFFICER AGENCY: FFWCC (West Palm)  
LAST AGENCY INSPECTION: 03/10/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1816818  
DRY GAS EXPIRY DATE: 02/13/2018  
VIOLATION CODE: BUI  
ARREST TIME: 17:30  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 18:45

SUBJECT'S LAST NAME: VECIANA

DATE/TIME OF TEST: 04/09/2016 19:16:01

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 19:08 |
| Air Blank          | 0.000 | 19:09 |
| Control Test       | 0.081 | 19:09 |
| Air Blank          | 0.000 | 19:09 |
| Subject Sample # 1 | 0.129 | 19:11 |
| Breath Volume      | 2.152 | 19:11 |
| Air Blank          | 0.000 | 19:12 |
| Air Blank          | 0.000 | 19:13 |
| Subject Sample # 2 | 0.140 | 19:14 |
| Breath Volume      | 2.000 | 19:14 |
| Air Blank          | 0.000 | 19:15 |
| Control Test       | 0.080 | 19:15 |
| Air Blank          | 0.000 | 19:15 |
| Diagnostics Check  | OK    | 19:16 |

---

Data Download Date/Time: 04/28/2016 09:19

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ELMORE,BENJAMIN,R  
BREATH TEST OPERATOR AGENCY: Monroe County SO  
ARRESTING OFFICER: ELMORE,BENJAMIN  
ARRESTING OFFICER AGENCY: Monroe County SO  
LAST AGENCY INSPECTION: 03/10/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1816818  
DRY GAS EXPIRY DATE: 02/13/2018  
VIOLATION CODE: DUI  
ARREST TIME: 07:16  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 07:31

SUBJECT'S LAST NAME: FALS MONTERO

DATE/TIME OF TEST: 04/10/2016 07:54:11

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 07:51 |                      |
| Air Blank         | 0.000 | 07:51 |                      |
| Control Test      | 0.081 | 07:52 |                      |
| Air Blank         | 0.000 | 07:52 |                      |
| Subject Sample    | # 1   | 07:52 | Subject Test Refused |
| Breath Volume     | 0.000 | 07:52 |                      |
| Air Blank         | 0.000 | 07:53 |                      |
| Control Test      | 0.081 | 07:53 |                      |
| Air Blank         | 0.000 | 07:54 |                      |
| Diagnostics Check | OK    | 07:54 |                      |

---

Data Download Date/Time: 04/28/2016 09:19

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TORRES,DEANNA,N  
BREATH TEST OPERATOR AGENCY: Monroe County SO  
ARRESTING OFFICER: ONEILL,MATTHEW  
ARRESTING OFFICER AGENCY: Monroe County SO  
LAST AGENCY INSPECTION: 03/10/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1816818  
DRY GAS EXPIRY DATE: 02/13/2018  
VIOLATION CODE: DUI  
ARREST TIME: 20:59  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:37

SUBJECT'S LAST NAME: ROUSSELL

DATE/TIME OF TEST: 04/10/2016 21:59:55

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 21:57 |                      |
| Air Blank         | 0.000 | 21:57 |                      |
| Control Test      | 0.082 | 21:57 |                      |
| Air Blank         | 0.000 | 21:58 |                      |
| Subject Sample    | # 1   | 21:58 | Subject Test Refused |
| Breath Volume     | 0.000 | 21:58 |                      |
| Air Blank         | 0.000 | 21:59 |                      |
| Control Test      | 0.081 | 21:59 |                      |
| Air Blank         | 0.000 | 21:59 |                      |
| Diagnostics Check | OK    | 21:59 |                      |

---

Data Download Date/Time: 04/28/2016 09:19

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TORRES,DEANNA,N  
BREATH TEST OPERATOR AGENCY: Monroe County SO  
ARRESTING OFFICER: ONSGARD,MICHAEL  
ARRESTING OFFICER AGENCY: Monroe County SO  
LAST AGENCY INSPECTION: 03/10/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1816818  
DRY GAS EXPIRY DATE: 02/13/2018  
VIOLATION CODE: DUI  
ARREST TIME: 21:48  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:34

SUBJECT'S LAST NAME: DIPERRI

DATE/TIME OF TEST: 04/10/2016 23:01:06

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:54 |
| Air Blank          | 0.000 | 22:55 |
| Control Test       | 0.081 | 22:55 |
| Air Blank          | 0.000 | 22:55 |
| Subject Sample # 1 | 0.093 | 22:56 |
| Breath Volume      | 1.203 | 22:56 |
| Air Blank          | 0.000 | 22:56 |
| Air Blank          | 0.000 | 22:58 |
| Subject Sample # 2 | 0.087 | 22:59 |
| Breath Volume      | 1.839 | 22:59 |
| Air Blank          | 0.000 | 23:00 |
| Control Test       | 0.081 | 23:00 |
| Air Blank          | 0.000 | 23:01 |
| Diagnostics Check  | OK    | 23:01 |

---

Data Download Date/Time: 04/28/2016 09:19

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KEITH,JASON,T  
BREATH TEST OPERATOR AGENCY: Monroe County SO  
ARRESTING OFFICER: KEITH,JASON  
ARRESTING OFFICER AGENCY: Monroe County SO  
LAST AGENCY INSPECTION: 03/10/2016

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 1816818  
DRY GAS EXPIRY DATE: 02/13/2018  
VIOLATION CODE: DUI  
ARREST TIME: 01:16  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: EGOICHEAGA

DATE/TIME OF TEST: 04/16/2016 02:35:39

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:29 |
| Air Blank          | 0.000 | 02:30 |
| Control Test       | 0.081 | 02:30 |
| Air Blank          | 0.000 | 02:30 |
| Subject Sample # 1 | 0.115 | 02:31 |
| Breath Volume      | 1.687 | 02:31 |
| Air Blank          | 0.000 | 02:31 |
| Air Blank          | 0.000 | 02:33 |
| Subject Sample # 2 | 0.115 | 02:34 |
| Breath Volume      | 2.027 | 02:34 |
| Air Blank          | 0.000 | 02:34 |
| Control Test       | 0.081 | 02:35 |
| Air Blank          | 0.000 | 02:35 |
| Diagnostics Check  | OK    | 02:35 |

---

Data Download Date/Time: 04/28/2016 09:19



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ORBETA,JORGE,  
BREATH TEST OPERATOR AGENCY: FHP Troop E  
ARRESTING OFFICER: ORBETA,JORGE  
ARRESTING OFFICER AGENCY: FHP Troop E  
LAST AGENCY INSPECTION: 03/10/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1816818  
DRY GAS EXPIRY DATE: 02/13/2018  
VIOLATION CODE: DUI  
ARREST TIME: 14:39  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 15:03

SUBJECT'S LAST NAME: ELIAS

DATE/TIME OF TEST: 04/16/2016 15:33:05

---

|                   |           |       |                 |
|-------------------|-----------|-------|-----------------|
| Diagnostics Check | OK        | 15:24 |                 |
| Air Blank         | 0.000     | 15:25 |                 |
| Control Test      | 0.081     | 15:25 |                 |
| Air Blank         | 0.000     | 15:26 |                 |
| Subject Sample    | # 1 0.243 | 15:29 | Volume Not Met  |
| Breath Volume     | 0.070     | 15:29 |                 |
| Air Blank         | 0.000     | 15:30 |                 |
| Air Blank         |           | 15:31 | Improper Sample |
| Air Blank         | 0.000     | 15:32 |                 |
| Control Test      | 0.081     | 15:32 |                 |
| Air Blank         | 0.000     | 15:32 |                 |
| Diagnostics Check | OK        | 15:33 |                 |

---

Data Download Date/Time: 04/28/2016 09:19

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ALVAREZ,ORLANDO,  
BREATH TEST OPERATOR AGENCY: Monroe County SO  
ARRESTING OFFICER: ALVAREZ,ORLANDO  
ARRESTING OFFICER AGENCY: Monroe County SO  
LAST AGENCY INSPECTION: 03/10/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1816818  
DRY GAS EXPIRY DATE: 02/13/2018  
VIOLATION CODE: DUI  
ARREST TIME: 01:43  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:07

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 04/24/2016 02:34:14

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:28 |
| Air Blank          | 0.000 | 02:28 |
| Control Test       | 0.081 | 02:28 |
| Air Blank          | 0.000 | 02:29 |
| Subject Sample # 1 | 0.147 | 02:29 |
| Breath Volume      | 1.718 | 02:29 |
| Air Blank          | 0.000 | 02:30 |
| Air Blank          | 0.000 | 02:32 |
| Subject Sample # 2 | 0.138 | 02:32 |
| Breath Volume      | 1.199 | 02:32 |
| Air Blank          | 0.000 | 02:33 |
| Control Test       | 0.081 | 02:33 |
| Air Blank          | 0.000 | 02:34 |
| Diagnostics Check  | OK    | 02:34 |

---

Data Download Date/Time: 04/28/2016 09:19

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ORBETA,JORGE,  
BREATH TEST OPERATOR AGENCY: FHP Troop E  
ARRESTING OFFICER: ORBETA,JORGE  
ARRESTING OFFICER AGENCY: FHP Troop E  
LAST AGENCY INSPECTION: 03/10/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1816818  
DRY GAS EXPIRY DATE: 02/13/2018  
VIOLATION CODE: DUI  
ARREST TIME: 18:33  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 18:52

SUBJECT'S LAST NAME: BENNETT

DATE/TIME OF TEST: 04/26/2016 19:15:27

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 19:12 |                      |
| Air Blank         | 0.000 | 19:13 |                      |
| Control Test      | 0.082 | 19:13 |                      |
| Air Blank         | 0.000 | 19:13 |                      |
| Subject Sample    | # 1   | 19:14 | Subject Test Refused |
| Breath Volume     | 0.000 | 19:14 |                      |
| Air Blank         | 0.000 | 19:14 |                      |
| Control Test      | 0.081 | 19:14 |                      |
| Air Blank         | 0.000 | 19:15 |                      |
| Diagnostics Check | OK    | 19:15 |                      |

---

Data Download Date/Time: 04/28/2016 09:19

# **Florida Department of Law Enforcement**

## **Alcohol Testing Program**

### **Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BRADY,JOSHUA,J  
BREATH TEST OPERATOR AGENCY: Monroe County SO  
ARRESTING OFFICER: HODGES,TITUS  
ARRESTING OFFICER AGENCY: Monroe County SO  
LAST AGENCY INSPECTION: 03/10/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1816818  
DRY GAS EXPIRY DATE: 02/13/2018  
VIOLATION CODE: DUI  
ARREST TIME: 20:34  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: DILLON

DATE/TIME OF TEST: 04/26/2016 21:23:17

---

|                   |       |       |            |
|-------------------|-------|-------|------------|
| Diagnostics Check | OK    | 21:21 |            |
| Air Blank         | 0.000 | 21:22 |            |
| Control Test      | 0.081 | 21:22 |            |
| Air Blank         |       | 21:22 | RFI Detect |
| Air Blank         | 0.000 | 21:23 |            |

---

Data Download Date/Time: 04/28/2016 09:19

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BRADY,JOSHUA,J  
BREATH TEST OPERATOR AGENCY: Monroe County SO  
ARRESTING OFFICER: HODGES,TITUS  
ARRESTING OFFICER AGENCY: Monroe County SO  
LAST AGENCY INSPECTION: 03/10/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1816818  
DRY GAS EXPIRY DATE: 02/13/2018  
VIOLATION CODE: DUI  
ARREST TIME: 20:34  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: DILLON

DATE/TIME OF TEST: 04/26/2016 21:32:45

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 21:26 |
| Air Blank          | 0.000 | 21:26 |
| Control Test       | 0.081 | 21:27 |
| Air Blank          | 0.000 | 21:27 |
| Subject Sample # 1 | 0.273 | 21:28 |
| Breath Volume      | 2.601 | 21:28 |
| Air Blank          | 0.000 | 21:28 |
| Air Blank          | 0.000 | 21:30 |
| Subject Sample # 2 | 0.255 | 21:31 |
| Breath Volume      | 3.253 | 21:31 |
| Air Blank          | 0.000 | 21:31 |
| Control Test       | 0.081 | 21:32 |
| Air Blank          | 0.000 | 21:32 |
| Diagnostics Check  | OK    | 21:32 |

---

Data Download Date/Time: 04/28/2016 09:19