Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 03/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31715080A3
DRY GAS EXPIRY DATE: 12/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:32

SUBJECT'S LAST NAME: ROSEMBERT

DATE/TIME OF TEST: 04/07/2016 03:59:48

Diagnostics Check		OK	03:53
Air Blank		0.000	03:54
Control Test		0.079	03:54
Air Blank		0.000	03:54
Subject Sample	# 1	0.021	03:55
Breath Volume		1.757	03:55
Air Blank		0.000	03:56
Air Blank		0.000	03:58
Subject Sample	# 2	0.019	03:58
Breath Volume		1.191	03:58
Air Blank		0.000	03:58
Control Test		0.079	03:59
Air Blank		0.000	03:59
Diagnostics Check		OK	03:59

Data Download Date/Time: 04/28/2016 15:02

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,RAUL,
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: GILPIN,SCHELISA
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 03/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1673300
DRY GAS EXPIRY DATE: 04/28/2017
VIOLATION CODE: DUI
ARREST TIME: 04:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: MONROE

DATE/TIME OF TEST: 04/16/2016 04:53:15

Diagnostics Check		OK	04:47	
Air Blank		0.000	04:47	
Control Test		0.080	04:48	
Air Blank		0.000	04:48	
Subject Sample	# 1		04:51	Subject Test Refused
Breath Volume		0.082	04:51	
Air Blank		0.000	04:52	
Control Test		0.081	04:52	
Air Blank		0.000	04:53	
Diagnostics Check		OK	04:53	

Data Download Date/Time: 04/28/2016 00:51

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,RAUL,
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: GILPIN,SCHELISA
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 03/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1673300
DRY GAS EXPIRY DATE: 04/28/2017
VIOLATION CODE: DUI
ARREST TIME: 03:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: MORALES RAMIREZ

DATE/TIME OF TEST: 04/17/2016 04:19:55

Diagnostics Check		OK	04:11	
Air Blank		0.000	04:11	
Control Test		0.081	04:12	
Air Blank		0.000	04:12	
Subject Sample	# 1		04:15	No Sample Provided
Breath Volume		0.000	04:15	
Air Blank		0.000	04:16	
Air Blank		0.000	04:18	
Subject Sample	# 2		04:18	Subject Test Refused
Breath Volume		0.000	04:18	
Air Blank		0.000	04:18	
Control Test		0.081	04:19	
Air Blank		0.000	04:19	
Diagnostics Check		OK	04:19	

Data Download Date/Time: 04/28/2016 00:51

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEPHENSON,JENNIFER,L
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: STEPHENSON,JENNIFER
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 03/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1673300
DRY GAS EXPIRY DATE: 04/28/2017
VIOLATION CODE: DUI
ARREST TIME: 05:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:12

SUBJECT'S LAST NAME: VELASQUEZ

DATE/TIME OF TEST: 04/17/2016 05:52:35

Diagnostics Check		OK	05:45
Air Blank		0.000	05:45
Control Test		0.081	05:46
Air Blank		0.000	05:46
Subject Sample	# 1	0.121	05:48
Breath Volume		1.464	05:48
Air Blank		0.000	05:48
Air Blank		0.000	05:50
Subject Sample	# 2	0.120	05:50
Breath Volume		1.886	05:50
Air Blank		0.000	05:51
Control Test		0.079	05:51
Air Blank		0.000	05:52
Diagnostics Check		OK	05:52

Data Download Date/Time: 04/28/2016 00:51

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NEWTON,JOSHUA,
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: NEWTON,JOSHUA
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 03/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1870563
DRY GAS EXPIRY DATE: 04/28/2017
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:38

SUBJECT'S LAST NAME: JACQUES

DATE/TIME OF TEST: 04/21/2016 03:13:26

Diagnostics Check		OK	03:06
Air Blank		0.000	03:06
Control Test		0.081	03:07
Air Blank		0.000	03:07
Subject Sample	# 1	0.205	03:08
Breath Volume		2.531	03:08
Air Blank		0.000	03:09
Air Blank		0.000	03:11
Subject Sample	# 2	0.197	03:11
Breath Volume		2.652	03:11
Air Blank		0.000	03:12
Control Test		0.080	03:12
Air Blank		0.000	03:13
Diagnostics Check		OK	03:13

Data Download Date/Time: 04/28/2016 00:51

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,RAUL,
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: PEREZ,RAUL
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 03/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1673300
DRY GAS EXPIRY DATE: 04/28/2017
VIOLATION CODE: DUI
ARREST TIME: 01:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: SILVA

DATE/TIME OF TEST: 04/23/2016 02:19:51

Diagnostics Check		OK	02:13
Air Blank		0.000	02:13
Control Test		0.080	02:14
Air Blank		0.000	02:14
Subject Sample	# 1	0.143	02:15
Breath Volume		1.906	02:15
Air Blank		0.000	02:16
Air Blank		0.000	02:17
Subject Sample	# 2	0.139	02:18
Breath Volume		2.500	02:18
Air Blank		0.000	02:18
Control Test		0.079	02:19
Air Blank		0.000	02:19
Diagnostics Check		OK	02:19

Data Download Date/Time: 04/28/2016 00:51

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEPHENSON,JENNIFER,L
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: STEPHENSON,JENNIFER
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 03/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1673300
DRY GAS EXPIRY DATE: 04/28/2017
VIOLATION CODE: DUI
ARREST TIME: 03:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:22

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 04/23/2016 04:04:09

Diagnostics Check		OK	04:01	
Air Blank		0.000	04:01	
Control Test		0.080	04:02	
Air Blank		0.000	04:02	
Subject Sample	# 1		04:02	Subject Test Refused
Breath Volume		0.000	04:02	
Air Blank		0.000	04:03	
Control Test		0.080	04:03	
Air Blank		0.000	04:04	
Diagnostics Check		OK	04:04	

Data Download Date/Time: 04/28/2016 00:51

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEPHENSON,JENNIFER,L
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: LUGO,VICTOR
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 03/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1673300
DRY GAS EXPIRY DATE: 04/28/2017
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:06

SUBJECT'S LAST NAME: TORRES

DATE/TIME OF TEST: 04/24/2016 01:41:39

Diagnostics Check		OK	01:38	
Air Blank		0.000	01:39	
Control Test		0.081	01:39	
Air Blank		0.000	01:40	
Subject Sample	# 1		01:40	Subject Test Refused
Breath Volume		0.000	01:40	
Air Blank		0.000	01:40	
Control Test		0.081	01:41	
Air Blank		0.000	01:41	
Diagnostics Check		OK	01:41	

Data Download Date/Time: 04/28/2016 00:51

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,JESUS,O
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: ACOSTA,MAYKEL
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 04/27/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1870563
DRY GAS EXPIRY DATE: 06/17/2018
VIOLATION CODE: DUI
ARREST TIME: 23:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:41

SUBJECT'S LAST NAME: MAY

DATE/TIME OF TEST: 04/28/2016 00:23:36

Diagnostics Check		OK	00:16
Air Blank		0.000	00:17
Control Test		0.081	00:17
Air Blank		0.000	00:18
Subject Sample	# 1	0.196	00:18
Breath Volume		2.980	00:18
Air Blank		0.000	00:19
Air Blank		0.000	00:21
Subject Sample	# 2	0.190	00:21
Breath Volume		2.738	00:21
Air Blank		0.000	00:22
Control Test		0.079	00:23
Air Blank		0.000	00:23
Diagnostics Check		OK	00:23

Data Download Date/Time: 04/28/2016 00:51

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,RAUL,
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: PEREZ,RAUL
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 04/27/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1870563
DRY GAS EXPIRY DATE: 06/17/2018
VIOLATION CODE: DUI
ARREST TIME: 01:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: DIAZ PENEDA

DATE/TIME OF TEST: 04/30/2016 02:01:18

Diagnostics Check		OK	01:54
Air Blank		0.000	01:54
Control Test		0.080	01:55
Air Blank		0.000	01:55
Subject Sample	# 1	0.210	01:56
Breath Volume		1.781	01:56
Air Blank		0.000	01:57
Air Blank		0.000	01:58
Subject Sample	# 2	0.218	01:59
Breath Volume		1.914	01:59
Air Blank		0.000	02:00
Control Test		0.080	02:00
Air Blank		0.000	02:01
Diagnostics Check		OK	02:01

Data Download Date/Time: 06/04/2016 10:18

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006214

INSTRUMENT REGISTERED TO: OCALA POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOEDICKE,WILLIAM,R
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: JOEDICKE,WILLIAM
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 03/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 19:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:10

SUBJECT'S LAST NAME: NELSON

DATE/TIME OF TEST: 04/09/2016 19:42:26

Diagnostics Check		OK	19:33
Air Blank		0.000	19:34
Control Test		0.080	19:34
Air Blank		0.000	19:34
Subject Sample	# 1	0.114	19:37
Breath Volume		2.167	19:37
Air Blank		0.000	19:38
Air Blank		0.000	19:39
Subject Sample	# 2	0.112	19:40
Breath Volume		1.203	19:40
Air Blank		0.000	19:41
Control Test		0.079	19:41
Air Blank		0.000	19:42
Diagnostics Check		OK	19:42

Data Download Date/Time: 04/27/2016 20:54

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006214

INSTRUMENT REGISTERED TO: OCALA POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,GARY,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: MILLER,GARY
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 03/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:58

SUBJECT'S LAST NAME: VELEZ MALDONADO

DATE/TIME OF TEST: 04/10/2016 03:21:06

Diagnostics Check		OK	03:18	
Air Blank		0.000	03:18	
Control Test		0.080	03:19	
Air Blank		0.000	03:19	
Subject Sample	# 1		03:19	RFI Detect
Breath Volume		0.000	03:19	
Air Blank		0.000	03:20	
Control Test		0.079	03:20	
Air Blank		0.000	03:21	
Diagnostics Check		OK	03:21	

Data Download Date/Time: 04/27/2016 20:54

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006214

INSTRUMENT REGISTERED TO: OCALA POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,GARY,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: MILLER,GARY
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 03/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:58

SUBJECT'S LAST NAME: VELEZ MALDONADO

DATE/TIME OF TEST: 04/10/2016 03:31:51

Diagnostics Check		OK	03:22
Air Blank		0.000	03:22
Control Test		0.079	03:23
Air Blank		0.000	03:23
Subject Sample	# 1	0.206	03:24
Breath Volume		1.472	03:24
Air Blank		0.000	03:25
Air Blank		0.000	03:26
Subject Sample	# 2	0.178	03:27
Breath Volume		1.257	03:27
Air Blank		0.000	03:27
Air Blank		0.000	03:29
Subject Sample	# 3	0.186	03:30
Breath Volume		1.964	03:30
Air Blank		0.000	03:30
Control Test		0.078	03:31
Air Blank		0.000	03:31
Diagnostics Check		OK	03:31

Data Download Date/Time: 04/27/2016 20:54

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006214

INSTRUMENT REGISTERED TO: OCALA POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOEDICKE,WILLIAM,R
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: JOEDICKE,WILLIAM
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 03/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:17

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 04/15/2016 02:50:28

Diagnostics Check		OK	02:41
Air Blank		0.000	02:41
Control Test		0.078	02:41
Air Blank		0.000	02:42
Subject Sample	# 1	0.183	02:42
Breath Volume		4.335	02:42
Air Blank		0.000	02:43
Air Blank		0.000	02:45
Subject Sample	# 2	0.155	02:45
Breath Volume		1.421	02:45
Air Blank		0.000	02:46
Air Blank		0.000	02:48
Subject Sample	# 3	0.182	02:48
Breath Volume		2.996	02:48
Air Blank		0.000	02:49
Control Test		0.078	02:49
Air Blank		0.000	02:50
Diagnostics Check		OK	02:50

Data Download Date/Time: 04/27/2016 20:54

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006214

INSTRUMENT REGISTERED TO: OCALA POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOEDICKE,WILLIAM,R
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: JOEDICKE,WILLIAM
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 03/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:53

SUBJECT'S LAST NAME: KENISTON

DATE/TIME OF TEST: 04/22/2016 03:20:47

Diagnostics Check		OK	03:14
Air Blank		0.000	03:15
Control Test		0.080	03:15
Air Blank		0.000	03:15
Subject Sample	# 1	0.116	03:16
Breath Volume		1.750	03:16
Air Blank		0.000	03:16
Air Blank		0.000	03:18
Subject Sample	# 2	0.113	03:19
Breath Volume		1.773	03:19
Air Blank		0.000	03:19
Control Test		0.078	03:20
Air Blank		0.000	03:20
Diagnostics Check		OK	03:20

Data Download Date/Time: 04/27/2016 20:54

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006214

INSTRUMENT REGISTERED TO: OCALA POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOEDICKE,WILLIAM,R
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: JOEDICKE,WILLIAM
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 04/27/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:53

SUBJECT'S LAST NAME: WOODS

DATE/TIME OF TEST: 04/28/2016 20:16:37

Diagnostics Check		OK	20:13	
Air Blank		0.000	20:14	
Control Test		0.077	20:14	
Air Blank		0.000	20:14	
Subject Sample	# 1		20:15	RFI Detect
Breath Volume		1.199	20:15	
Air Blank		0.000	20:15	
Control Test		0.076	20:16	
Air Blank		0.000	20:16	
Diagnostics Check		OK	20:16	

Data Download Date/Time: 05/18/2016 21:02

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006214

INSTRUMENT REGISTERED TO: OCALA POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOEDICKE,WILLIAM,R
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: JOEDICKE,WILLIAM
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 04/27/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:53

SUBJECT'S LAST NAME: WOODS

DATE/TIME OF TEST: 04/28/2016 20:24:43

Diagnostics Check		OK	20:18
Air Blank		0.000	20:18
Control Test		0.077	20:19
Air Blank		0.000	20:19
Subject Sample	# 1	0.123	20:20
Breath Volume		2.339	20:20
Air Blank		0.000	20:20
Air Blank		0.000	20:22
Subject Sample	# 2	0.124	20:23
Breath Volume		2.390	20:23
Air Blank		0.000	20:23
Control Test		0.077	20:24
Air Blank		0.000	20:24
Diagnostics Check		OK	20:24

Data Download Date/Time: 05/18/2016 21:02

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006222

INSTRUMENT REGISTERED TO: ASTATULA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,RODNEY,A
BREATH TEST OPERATOR AGENCY: Astatula PD
ARRESTING OFFICER: REVELT,GARY
ARRESTING OFFICER AGENCY: Groveland PD
LAST AGENCY INSPECTION: 03/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09014080A1
DRY GAS EXPIRY DATE: 05/01/2016
VIOLATION CODE: DUI
ARREST TIME: 00:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: MAHON

DATE/TIME OF TEST: 04/04/2016 01:52:04

Diagnostics Check		OK	01:49	
Air Blank		0.000	01:49	
Control Test		0.079	01:49	
Air Blank		0.000	01:50	
Subject Sample	# 1		01:50	Subject Test Refused
Breath Volume		0.000	01:50	
Air Blank		0.000	01:51	
Control Test		0.079	01:51	
Air Blank		0.000	01:51	
Diagnostics Check		OK	01:52	

Data Download Date/Time: 04/30/2016 21:33

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAGNER,ROBERT,F
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: WAGNER,ROBERT
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 03/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG425202
DRY GAS EXPIRY DATE: 09/09/2016
VIOLATION CODE: DUI
ARREST TIME: 02:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:14

SUBJECT'S LAST NAME: WISHEART

DATE/TIME OF TEST: 04/07/2016 02:44:11

Diagnostics Check		OK	02:37
Air Blank		0.000	02:38
Control Test		0.076	02:38
Air Blank		0.000	02:39
Subject Sample	# 1	0.178	02:39
Breath Volume		3.156	02:39
Air Blank		0.000	02:40
Air Blank		0.000	02:42
Subject Sample	# 2	0.180	02:42
Breath Volume		2.023	02:42
Air Blank		0.000	02:43
Control Test		0.075	02:43
Air Blank		0.000	02:44
Diagnostics Check		OK	02:44

Data Download Date/Time: 04/28/2016 12:35

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASE,THOMAS,D
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: HART,JEREMY
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 03/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG425202
DRY GAS EXPIRY DATE: 09/09/2016
VIOLATION CODE: DUI
ARREST TIME: 01:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:11

SUBJECT'S LAST NAME: JERNIGAN

DATE/TIME OF TEST: 04/18/2016 02:31:02

Diagnostics Check		OK	02:23
Air Blank		0.000	02:23
Control Test		0.078	02:24
Air Blank		0.000	02:24
Subject Sample	# 1	0.000	02:26
Breath Volume		1.250	02:26
Air Blank		0.000	02:26
Air Blank		0.000	02:28
Subject Sample	# 2	0.000	02:29
Breath Volume		2.777	02:29
Air Blank		0.000	02:30
Control Test		0.077	02:30
Air Blank		0.000	02:30
Diagnostics Check		OK	02:31

Data Download Date/Time: 04/28/2016 12:35

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASE,THOMAS,D
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: BIVINS,HEATHER
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 03/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG425202
DRY GAS EXPIRY DATE: 09/09/2016
VIOLATION CODE: DUI
ARREST TIME: 02:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:16

SUBJECT'S LAST NAME: GLADMAN

DATE/TIME OF TEST: 04/18/2016 02:50:56

Diagnostics Check		OK	02:42	
Air Blank		0.000	02:42	
Control Test		0.077	02:43	
Air Blank		0.000	02:43	
Subject Sample	# 1	0.213	02:46	
Breath Volume		1.386	02:46	
Air Blank		0.000	02:46	
Air Blank		0.000	02:48	
Subject Sample	# 2	0.212	02:49	
Breath Volume		1.253	02:49	
Air Blank		0.000	02:50	
Control Test		0.074	02:50	Control Outside Tolerance
Air Blank		0.000	02:50	

Data Download Date/Time: 04/28/2016 12:35

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARRITY,THOMAS,A
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: GLEASON,WILLIAM
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 03/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG425202
DRY GAS EXPIRY DATE: 09/09/2016
VIOLATION CODE: DUI
ARREST TIME: 20:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:53

SUBJECT'S LAST NAME: GUENTHER

DATE/TIME OF TEST: 04/18/2016 21:25:22

Diagnostics Check		OK	21:13	
Air Blank		0.000	21:13	
Control Test		0.077	21:14	
Air Blank		0.000	21:14	
Subject Sample	# 1	0.144	21:17	Volume Not Met
Breath Volume		0.515	21:17	
Air Blank		0.000	21:18	
Air Blank		0.000	21:20	
Subject Sample	# 2	0.172	21:20	
Breath Volume		1.199	21:20	
Air Blank		0.000	21:21	
Air Blank		0.000	21:23	
Subject Sample	# 3	0.169	21:23	
Breath Volume		1.117	21:23	
Air Blank		0.000	21:24	
Control Test		0.075	21:24	
Air Blank		0.000	21:25	
Diagnostics Check		OK	21:25	

Data Download Date/Time: 04/28/2016 12:35

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASE,THOMAS,D
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: ROWELL,ROBERT
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 04/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG425202
DRY GAS EXPIRY DATE: 09/09/2016
VIOLATION CODE: DUI
ARREST TIME: 00:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:08

SUBJECT'S LAST NAME: ROBERTS

DATE/TIME OF TEST: 04/30/2016 00:40:24

Diagnostics Check		OK	00:31
Air Blank		0.000	00:32
Control Test		0.077	00:32
Air Blank		0.000	00:32
Subject Sample	# 1	0.064	00:35
Breath Volume		2.972	00:35
Air Blank		0.000	00:36
Air Blank		0.000	00:38
Subject Sample	# 2	0.065	00:38
Breath Volume		2.593	00:38
Air Blank		0.000	00:39
Control Test		0.077	00:39
Air Blank		0.000	00:40
Diagnostics Check		OK	00:40

Data Download Date/Time: 05/16/2016 10:27

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006234

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,ALFONSO,
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: RODRIGUEZ,ALFONSO
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 03/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1694353
DRY GAS EXPIRY DATE: 06/10/2017
VIOLATION CODE: DUI
ARREST TIME: 13:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:08

SUBJECT'S LAST NAME: BELLO HERNANDEZ

DATE/TIME OF TEST: 04/01/2016 16:52:53

Diagnostics Check		OK	16:38	
Air Blank		0.000	16:38	
Control Test		0.081	16:38	
Air Blank		0.000	16:39	
Subject Sample	# 1	0.202	16:42	Volume Not Met
Breath Volume		0.582	16:42	
Air Blank		0.000	16:43	
Air Blank		0.000	16:45	
Subject Sample	# 2	0.222	16:45	
Breath Volume		2.609	16:45	
Air Blank		0.000	16:46	
Air Blank		0.000	16:48	
Subject Sample	# 3	0.198	16:51	Volume Not Met
Breath Volume		1.195	16:51	
Air Blank		0.000	16:51	
Control Test		0.080	16:52	
Air Blank		0.000	16:52	
Diagnostics Check		OK	16:52	

Data Download Date/Time: 04/28/2016 14:23

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006234

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAYLOR,MICHAEL,R
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: NORWOOD,C
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 03/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1694353
DRY GAS EXPIRY DATE: 06/10/2017
VIOLATION CODE: DUI
ARREST TIME: 00:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:48

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 04/19/2016 01:26:21

Diagnostics Check		OK	01:20
Air Blank		0.000	01:20
Control Test		0.082	01:21
Air Blank		0.000	01:21
Subject Sample	# 1	0.146	01:21
Breath Volume		1.429	01:21
Air Blank		0.000	01:22
Air Blank		0.000	01:24
Subject Sample	# 2	0.151	01:24
Breath Volume		1.804	01:24
Air Blank		0.000	01:25
Control Test		0.077	01:25
Air Blank		0.000	01:26
Diagnostics Check		OK	01:26

Data Download Date/Time: 04/28/2016 14:23

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006234

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOODS,JOSHUA,H
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: BAYLY,B
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 03/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1694353
DRY GAS EXPIRY DATE: 06/10/2017
VIOLATION CODE: DUI
ARREST TIME: 16:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:20

SUBJECT'S LAST NAME: STUDLEY

DATE/TIME OF TEST: 04/19/2016 17:09:48

Diagnostics Check		OK	17:06	
Air Blank		0.000	17:07	
Control Test		0.081	17:07	
Air Blank		0.000	17:08	
Subject Sample	# 1		17:08	Subject Test Refused
Breath Volume		0.000	17:08	
Air Blank		0.000	17:08	
Control Test		0.081	17:09	
Air Blank		0.000	17:09	
Diagnostics Check		OK	17:09	

Data Download Date/Time: 04/28/2016 14:23

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZIER,TAMMY,L
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: PRITCHARD,TRAVIS
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 03/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12014080A1
DRY GAS EXPIRY DATE: 06/01/2016
VIOLATION CODE: DUI
ARREST TIME: 21:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: GALLACHER

DATE/TIME OF TEST: 04/13/2016 22:30:50

Diagnostics Check		OK	22:22
Air Blank		0.000	22:23
Control Test		0.080	22:23
Air Blank		0.000	22:24
Subject Sample	# 1	0.206	22:24
Breath Volume		1.437	22:24
Air Blank		0.000	22:25
Air Blank		0.000	22:27
Subject Sample	# 2	0.209	22:29
Breath Volume		1.917	22:29
Air Blank		0.000	22:29
Control Test		0.080	22:30
Air Blank		0.000	22:30
Diagnostics Check		OK	22:30

Data Download Date/Time: 04/27/2016 13:54

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZIER,TAMMY,L
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: PRITCHARD,TRAVIS
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 03/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12014080A1
DRY GAS EXPIRY DATE: 06/01/2016
VIOLATION CODE: DUI
ARREST TIME: 23:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:53

SUBJECT'S LAST NAME: SIMPSON

DATE/TIME OF TEST: 04/15/2016 00:28:16

Diagnostics Check		OK	00:18
Air Blank		0.000	00:18
Control Test		0.079	00:19
Air Blank		0.000	00:19
Subject Sample	# 1	0.300	00:20
Breath Volume		1.773	00:20
Air Blank		0.000	00:21
Air Blank		0.000	00:22
Subject Sample	# 2	0.276	00:23
Breath Volume		1.679	00:23
Air Blank		0.000	00:24
Air Blank		0.000	00:26
Subject Sample	# 3	0.281	00:26
Breath Volume		1.453	00:26
Air Blank		0.000	00:27
Control Test		0.078	00:27
Air Blank		0.000	00:28
Diagnostics Check		OK	00:28

Data Download Date/Time: 04/27/2016 13:54

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STACY,A,L
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: SAUNDERS,J
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 03/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12014080A1
DRY GAS EXPIRY DATE: 06/01/2016
VIOLATION CODE: DUI
ARREST TIME: 20:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: HATCHER

DATE/TIME OF TEST: 04/15/2016 21:18:20

Diagnostics Check		OK	21:11
Air Blank		0.000	21:11
Control Test		0.080	21:12
Air Blank		0.000	21:12
Subject Sample	# 1	0.166	21:13
Breath Volume		1.687	21:13
Air Blank		0.000	21:13
Air Blank		0.000	21:15
Subject Sample	# 2	0.174	21:16
Breath Volume		1.437	21:16
Air Blank		0.000	21:17
Control Test		0.079	21:17
Air Blank		0.000	21:18
Diagnostics Check		OK	21:18

Data Download Date/Time: 04/27/2016 13:54

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STACY,A,L
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: SAUNDERS,J
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 03/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12014080A1
DRY GAS EXPIRY DATE: 06/01/2016
VIOLATION CODE: DUI
ARREST TIME: 21:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:12

SUBJECT'S LAST NAME: BECKWORTH

DATE/TIME OF TEST: 04/16/2016 22:40:07

Diagnostics Check		OK	22:33
Air Blank		0.000	22:33
Control Test		0.080	22:34
Air Blank		0.000	22:34
Subject Sample	# 1	0.170	22:35
Breath Volume		2.777	22:35
Air Blank		0.000	22:36
Air Blank		0.000	22:37
Subject Sample	# 2	0.160	22:38
Breath Volume		1.722	22:38
Air Blank		0.000	22:39
Control Test		0.079	22:39
Air Blank		0.000	22:40
Diagnostics Check		OK	22:40

Data Download Date/Time: 04/27/2016 13:54

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,JOHN,K
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: SPAULDING,T
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 03/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12014080A1
DRY GAS EXPIRY DATE: 06/01/2016
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: BLAKEWOOD

DATE/TIME OF TEST: 04/17/2016 03:37:23

Diagnostics Check		OK	03:30
Air Blank		0.000	03:30
Control Test		0.079	03:31
Air Blank		0.000	03:31
Subject Sample	# 1	0.000	03:33
Breath Volume		2.738	03:33
Air Blank		0.000	03:33
Air Blank		0.000	03:35
Subject Sample	# 2	0.000	03:36
Breath Volume		2.066	03:36
Air Blank		0.000	03:36
Control Test		0.079	03:36
Air Blank		0.000	03:37
Diagnostics Check		OK	03:37

Data Download Date/Time: 04/27/2016 13:54

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ASHE,KEVIN,B
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: SATTERWHITE,J
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 03/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12014080A1
DRY GAS EXPIRY DATE: 06/01/2016
VIOLATION CODE: DUI
ARREST TIME: 16:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:40

SUBJECT'S LAST NAME: HIRED

DATE/TIME OF TEST: 04/23/2016 18:19:29

Diagnostics Check		OK	18:08	
Air Blank		0.000	18:08	
Control Test		0.080	18:08	
Air Blank		0.000	18:09	
Subject Sample	# 1		18:12	No Sample Provided
Breath Volume		0.000	18:12	
Air Blank		0.000	18:12	
Air Blank		0.000	18:14	
Subject Sample	# 2		18:18	No Sample Provided
Breath Volume		0.000	18:18	
Air Blank		0.000	18:18	
Control Test		0.080	18:18	
Air Blank		0.000	18:19	
Diagnostics Check		OK	18:19	

Data Download Date/Time: 04/27/2016 13:54

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,JOHN,K
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: SAUNDERS,J
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 03/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12014080A1
DRY GAS EXPIRY DATE: 06/01/2016
VIOLATION CODE: DUI
ARREST TIME: 00:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:16

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 04/26/2016 01:49:57

Diagnostics Check		OK	01:43
Air Blank		0.000	01:44
Control Test		0.080	01:44
Air Blank		0.000	01:44
Subject Sample	# 1	0.201	01:45
Breath Volume		1.367	01:45
Air Blank		0.000	01:46
Air Blank		0.000	01:47
Subject Sample	# 2	0.199	01:48
Breath Volume		3.394	01:48
Air Blank		0.000	01:49
Control Test		0.079	01:49
Air Blank		0.000	01:49
Diagnostics Check		OK	01:49

Data Download Date/Time: 04/27/2016 13:54

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,JOHN,K
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: SAUNDERS,J
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 03/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12014080A1
DRY GAS EXPIRY DATE: 06/01/2016
VIOLATION CODE: DUI
ARREST TIME: 17:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:40

SUBJECT'S LAST NAME: WARREN

DATE/TIME OF TEST: 04/26/2016 19:08:45

Diagnostics Check		OK	19:02
Air Blank		0.000	19:02
Control Test		0.080	19:02
Air Blank		0.000	19:03
Subject Sample	# 1	0.278	19:03
Breath Volume		3.601	19:03
Air Blank		0.000	19:04
Air Blank		0.000	19:06
Subject Sample	# 2	0.279	19:07
Breath Volume		3.800	19:07
Air Blank		0.000	19:07
Control Test		0.080	19:08
Air Blank		0.000	19:08
Diagnostics Check		OK	19:08

Data Download Date/Time: 04/27/2016 13:54

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006262

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YACKEL,NICHOLAS,M
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: TATUM,THOMAS
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 03/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 23:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:33

SUBJECT'S LAST NAME: MCMASTER

DATE/TIME OF TEST: 04/16/2016 00:00:10

Diagnostics Check		OK	23:53	
Air Blank		0.000	23:54	
Control Test		0.081	23:54	
Air Blank		0.000	23:54	
Subject Sample	# 1	0.039	23:55	
Breath Volume		1.500	23:55	
Air Blank		0.000	23:56	
Air Blank		0.000	23:57	
Subject Sample	# 2	0.033	23:58	
Breath Volume		1.671	23:58	
Air Blank		0.000	23:59	
Control Test		0.069	23:59	RFI Detect
Air Blank		0.000	00:00	

Data Download Date/Time: 04/19/2016 15:46

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006262

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YACKEL,NICHOLAS,M
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: TATUM,THOMAS
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 03/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 23:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:33

SUBJECT'S LAST NAME: MCMASTER

DATE/TIME OF TEST: 04/16/2016 00:38:03

Diagnostics Check		OK	00:25	
Air Blank		0.000	00:25	
Control Test		0.081	00:25	
Air Blank		0.000	00:26	
Subject Sample	# 1	0.028	00:27	
Breath Volume		1.441	00:27	
Air Blank		0.000	00:27	
Air Blank		0.000	00:29	
Subject Sample	# 2	0.022	00:32	Volume Not Met
Breath Volume		0.222	00:32	
Air Blank		0.000	00:33	
Air Blank		0.000	00:35	
Subject Sample	# 3		00:36	Subject Test Refused
Breath Volume		0.066	00:36	
Air Blank		0.000	00:37	
Control Test		0.077	00:37	
Air Blank		0.000	00:37	
Diagnostics Check		OK	00:38	

Data Download Date/Time: 04/19/2016 15:46

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006262

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRYAN,STEVEN,A
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: FROST,JOSEPH
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 03/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 00:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: BAKER

DATE/TIME OF TEST: 04/16/2016 01:02:20

Diagnostics Check		OK	00:55
Air Blank		0.000	00:56
Control Test		0.079	00:56
Air Blank		0.000	00:57
Subject Sample	# 1	0.155	00:57
Breath Volume		1.738	00:57
Air Blank		0.000	00:58
Air Blank		0.000	01:00
Subject Sample	# 2	0.151	01:00
Breath Volume		2.515	01:00
Air Blank		0.000	01:01
Control Test		0.079	01:01
Air Blank		0.000	01:02
Diagnostics Check		OK	01:02

Data Download Date/Time: 04/19/2016 15:46

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006262

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: IHNKEN,MATTHEW,D
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: IHNKEN,MATTHEW
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 03/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 18:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:45

SUBJECT'S LAST NAME: CROUSE

DATE/TIME OF TEST: 04/16/2016 19:13:09

Diagnostics Check		OK	19:06
Air Blank		0.000	19:07
Control Test		0.080	19:07
Air Blank		0.000	19:08
Subject Sample	# 1	0.205	19:08
Breath Volume		1.582	19:08
Air Blank		0.000	19:09
Air Blank		0.000	19:11
Subject Sample	# 2	0.209	19:11
Breath Volume		1.476	19:11
Air Blank		0.000	19:12
Control Test		0.080	19:12
Air Blank		0.000	19:13
Diagnostics Check		OK	19:13

Data Download Date/Time: 04/19/2016 15:46

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006262

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GORANOVICH,AMY,L
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: SELTZER,MATTHEW
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 03/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 22:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:05

SUBJECT'S LAST NAME: CRUZ

DATE/TIME OF TEST: 04/16/2016 23:33:00

Diagnostics Check		OK	23:25
Air Blank		0.000	23:26
Control Test		0.081	23:26
Air Blank		0.000	23:27
Subject Sample	# 1	0.198	23:28
Breath Volume		1.476	23:28
Air Blank		0.000	23:29
Air Blank		0.000	23:30
Subject Sample	# 2	0.214	23:31
Breath Volume		1.632	23:31
Air Blank		0.000	23:32
Control Test		0.080	23:32
Air Blank		0.000	23:32
Diagnostics Check		OK	23:33

Data Download Date/Time: 04/19/2016 15:46

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006262

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PECK,KEITH,D
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: PECK,KEITH
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 04/19/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 03:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:48

SUBJECT'S LAST NAME: MCLOUGHLIN

DATE/TIME OF TEST: 04/26/2016 04:15:44

Diagnostics Check		OK	04:09
Air Blank		0.000	04:09
Control Test		0.081	04:09
Air Blank		0.000	04:10
Subject Sample	# 1	0.133	04:10
Breath Volume		2.828	04:10
Air Blank		0.000	04:11
Air Blank		0.000	04:13
Subject Sample	# 2	0.123	04:14
Breath Volume		1.914	04:14
Air Blank		0.000	04:14
Control Test		0.080	04:15
Air Blank		0.000	04:15
Diagnostics Check		OK	04:15

Data Download Date/Time: 05/05/2016 11:26

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006384

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEMEN,AARON,M
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: STEMEN,AARON
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 03/16/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:37

SUBJECT'S LAST NAME: KECK

DATE/TIME OF TEST: 04/02/2016 02:10:23

Diagnostics Check		OK	01:59
Air Blank		0.000	01:59
Control Test		0.079	02:00
Air Blank		0.000	02:00
Subject Sample	# 1	0.264	02:01
Breath Volume		1.574	02:01
Air Blank		0.000	02:02
Air Blank		0.000	02:04
Subject Sample	# 2	0.288	02:05
Breath Volume		2.140	02:05
Air Blank		0.000	02:06
Air Blank		0.000	02:08
Subject Sample	# 3	0.280	02:08
Breath Volume		1.750	02:08
Air Blank		0.000	02:09
Control Test		0.077	02:09
Air Blank		0.000	02:10
Diagnostics Check		OK	02:10

Data Download Date/Time: 04/21/2016 21:22

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006384

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SOWARDS,JACOB,T
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: WELCH,STEPHANIE
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 03/16/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:54

SUBJECT'S LAST NAME: LE

DATE/TIME OF TEST: 04/03/2016 04:26:22

Diagnostics Check		OK	04:14	
Air Blank		0.000	04:15	
Control Test		0.081	04:15	
Air Blank		0.000	04:16	
Subject Sample	# 1		04:19	No Sample Provided
Breath Volume		0.000	04:19	
Air Blank		0.000	04:19	
Air Blank		0.000	04:21	
Subject Sample	# 2		04:24	No Sample Provided
Breath Volume		0.000	04:24	
Air Blank		0.000	04:25	
Control Test		0.081	04:25	
Air Blank		0.000	04:26	
Diagnostics Check		OK	04:26	

Data Download Date/Time: 04/21/2016 21:22

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006384

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEMEN,AARON,M
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: STEMEN,AARON
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 03/16/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:36

SUBJECT'S LAST NAME: SHEA

DATE/TIME OF TEST: 04/08/2016 04:06:11

Diagnostics Check		OK	03:58
Air Blank		0.000	03:58
Control Test		0.080	03:59
Air Blank		0.000	03:59
Subject Sample	# 1	0.195	04:00
Breath Volume		1.796	04:00
Air Blank		0.000	04:01
Air Blank		0.000	04:02
Subject Sample	# 2	0.186	04:04
Breath Volume		1.828	04:04
Air Blank		0.000	04:05
Control Test		0.080	04:05
Air Blank		0.000	04:06
Diagnostics Check		OK	04:06

Data Download Date/Time: 04/21/2016 21:22

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006384

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SOWARDS,JACOB,T
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: BRINSKO,THOMAS
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 03/16/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: ENGLISH

DATE/TIME OF TEST: 04/08/2016 04:38:36

Diagnostics Check		OK	04:31
Air Blank		0.000	04:32
Control Test		0.079	04:32
Air Blank		0.000	04:32
Subject Sample	# 1	0.235	04:33
Breath Volume		1.582	04:33
Air Blank		0.000	04:34
Air Blank		0.000	04:36
Subject Sample	# 2	0.229	04:36
Breath Volume		1.589	04:36
Air Blank		0.000	04:37
Control Test		0.076	04:37
Air Blank		0.000	04:38
Diagnostics Check		OK	04:38

Data Download Date/Time: 04/21/2016 21:22

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006384

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEMEN,AARON,M
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: STEMEN,AARON
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 03/16/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:36

SUBJECT'S LAST NAME: JUDGE

DATE/TIME OF TEST: 04/09/2016 01:09:27

Diagnostics Check		OK	01:00	
Air Blank		0.000	01:01	
Control Test		0.080	01:01	
Air Blank		0.000	01:01	
Subject Sample	# 1		01:05	No Sample Provided
Breath Volume		0.000	01:05	
Air Blank		0.000	01:05	
Air Blank		0.000	01:07	
Subject Sample	# 2		01:08	Subject Test Refused
Breath Volume		0.000	01:08	
Air Blank		0.000	01:08	
Control Test		0.081	01:08	
Air Blank		0.000	01:09	
Diagnostics Check		OK	01:09	

Data Download Date/Time: 04/21/2016 21:22

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006384

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,THEODORE,D
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: BRINSKO,THOMAS
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 03/16/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: BURNETT

DATE/TIME OF TEST: 04/09/2016 02:12:00

Diagnostics Check		OK	02:05
Air Blank		0.000	02:05
Control Test		0.080	02:06
Air Blank		0.000	02:06
Subject Sample	# 1	0.145	02:07
Breath Volume		1.652	02:07
Air Blank		0.000	02:07
Air Blank		0.000	02:09
Subject Sample	# 2	0.127	02:10
Breath Volume		1.632	02:10
Air Blank		0.000	02:11
Control Test		0.078	02:11
Air Blank		0.000	02:11
Diagnostics Check		OK	02:12

Data Download Date/Time: 04/21/2016 21:22

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006384

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEMEN,AARON,M
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: STEMEN,AARON
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 03/16/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:59

SUBJECT'S LAST NAME: RAMIREZ

DATE/TIME OF TEST: 04/10/2016 02:30:31

Diagnostics Check		OK	02:23
Air Blank		0.000	02:23
Control Test		0.080	02:24
Air Blank		0.000	02:24
Subject Sample	# 1	0.161	02:26
Breath Volume		1.652	02:26
Air Blank		0.000	02:26
Air Blank		0.000	02:28
Subject Sample	# 2	0.174	02:28
Breath Volume		2.117	02:28
Air Blank		0.000	02:29
Control Test		0.079	02:29
Air Blank		0.000	02:30
Diagnostics Check		OK	02:30

Data Download Date/Time: 04/21/2016 21:22

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006384

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODONIS,AMANDA,J
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: BRICE,DAVID
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 03/16/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 15:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:18

SUBJECT'S LAST NAME: ACOSTA

DATE/TIME OF TEST: 04/16/2016 16:45:53

Diagnostics Check		OK	16:39
Air Blank		0.000	16:39
Control Test		0.081	16:40
Air Blank		0.000	16:40
Subject Sample	# 1	0.230	16:41
Breath Volume		3.050	16:41
Air Blank		0.000	16:41
Air Blank		0.000	16:43
Subject Sample	# 2	0.243	16:44
Breath Volume		2.714	16:44
Air Blank		0.000	16:44
Control Test		0.077	16:45
Air Blank		0.000	16:45
Diagnostics Check		OK	16:45

Data Download Date/Time: 04/21/2016 21:22

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006384

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WELCH,CHAD,J
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: WELCH,CHAD
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 03/16/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: WALDOR

DATE/TIME OF TEST: 04/18/2016 03:03:05

Diagnostics Check		OK	02:56
Air Blank		0.000	02:56
Control Test		0.080	02:57
Air Blank		0.000	02:57
Subject Sample	# 1	0.228	02:58
Breath Volume		4.144	02:58
Air Blank		0.000	02:59
Air Blank		0.000	03:00
Subject Sample	# 2	0.218	03:01
Breath Volume		3.292	03:01
Air Blank		0.000	03:02
Control Test		0.077	03:02
Air Blank		0.000	03:03
Diagnostics Check		OK	03:03

Data Download Date/Time: 04/21/2016 21:22

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006384

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROWE,JASON,S
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: ROWE,JASON
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 03/16/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 11:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:30

SUBJECT'S LAST NAME: CADE

DATE/TIME OF TEST: 04/20/2016 11:47:51

Diagnostics Check		OK	11:41
Air Blank		0.000	11:42
Control Test		0.080	11:42
Air Blank		0.000	11:43
Subject Sample	# 1	0.000	11:43
Breath Volume		4.171	11:43
Air Blank		0.000	11:43
Air Blank		0.000	11:45
Subject Sample	# 2	0.000	11:46
Breath Volume		2.843	11:46
Air Blank		0.000	11:46
Control Test		0.077	11:47
Air Blank		0.000	11:47
Diagnostics Check		OK	11:47

Data Download Date/Time: 04/21/2016 21:22

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006384

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEMEN,AARON,M
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: STEMEN,AARON
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 04/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:51

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 04/22/2016 01:19:56

Diagnostics Check		OK	01:12
Air Blank		0.000	01:13
Control Test		0.080	01:13
Air Blank		0.000	01:14
Subject Sample	# 1	0.061	01:15
Breath Volume		1.359	01:15
Air Blank		0.000	01:15
Air Blank		0.000	01:17
Subject Sample	# 2	0.058	01:18
Breath Volume		1.484	01:18
Air Blank		0.000	01:18
Control Test		0.080	01:19
Air Blank		0.000	01:19
Diagnostics Check		OK	01:19

Data Download Date/Time: 05/25/2016 19:43

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006384

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODONIS,AMANDA,J
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: RODONIS,AMANDA
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 04/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 09:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:10

SUBJECT'S LAST NAME: CALDWELL

DATE/TIME OF TEST: 04/24/2016 10:39:08

Diagnostics Check		OK	10:30
Air Blank		0.000	10:31
Control Test		0.080	10:31
Air Blank		0.000	10:31
Subject Sample	# 1	0.148	10:33
Breath Volume		1.203	10:33
Air Blank		0.000	10:34
Air Blank		0.000	10:35
Subject Sample	# 2	0.155	10:37
Breath Volume		1.304	10:37
Air Blank		0.000	10:38
Control Test		0.076	10:38
Air Blank		0.000	10:39
Diagnostics Check		OK	10:39

Data Download Date/Time: 05/25/2016 19:43

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006384

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WELCH,CHAD,J
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: WELCH,CHAD
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 04/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 19:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:35

SUBJECT'S LAST NAME: SHELLEY JR

DATE/TIME OF TEST: 04/24/2016 20:34:43

Diagnostics Check		OK	20:23	
Air Blank		0.000	20:23	
Control Test		0.079	20:23	
Air Blank		0.000	20:24	
Subject Sample	# 1		20:27	No Sample Provided
Breath Volume		0.000	20:27	
Air Blank		0.000	20:28	
Air Blank		0.000	20:30	
Subject Sample	# 2		20:33	No Sample Provided
Breath Volume		0.000	20:33	
Air Blank		0.000	20:33	
Control Test		0.080	20:34	
Air Blank		0.000	20:34	
Diagnostics Check		OK	20:34	

Data Download Date/Time: 05/25/2016 19:43

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006384

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALTERS,MATTHEW,N
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: WALTERS,MATTHEW
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 04/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 17:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:53

SUBJECT'S LAST NAME: BIERMAN

DATE/TIME OF TEST: 04/28/2016 18:21:34

Diagnostics Check		OK	18:14
Air Blank		0.000	18:15
Control Test		0.080	18:15
Air Blank		0.000	18:15
Subject Sample	# 1	0.343	18:16
Breath Volume		2.058	18:16
Air Blank		0.000	18:17
Air Blank		0.000	18:19
Subject Sample	# 2	0.338	18:19
Breath Volume		2.277	18:19
Air Blank		0.000	18:20
Control Test		0.078	18:20
Air Blank		0.000	18:21
Diagnostics Check		OK	18:21

Data Download Date/Time: 05/25/2016 19:43

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006384

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALTERS,MATTHEW,N
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: WALTERS,MATTHEW
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 04/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 14:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:07

SUBJECT'S LAST NAME: THORNTON

DATE/TIME OF TEST: 04/29/2016 15:31:09

Diagnostics Check		OK	15:28	
Air Blank		0.000	15:28	
Control Test		0.080	15:28	
Air Blank		0.000	15:29	
Subject Sample	# 1		15:29	Subject Test Refused
Breath Volume		0.000	15:29	
Air Blank		0.000	15:30	
Control Test		0.079	15:30	
Air Blank		0.000	15:31	
Diagnostics Check		OK	15:31	

Data Download Date/Time: 05/25/2016 19:43

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DE JESUS,V
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:19

SUBJECT'S LAST NAME: JUAN-MENESES

DATE/TIME OF TEST: 04/06/2016 23:52:28

Diagnostics Check		OK	23:42
Air Blank		0.000	23:43
Control Test		0.079	23:43
Air Blank		0.000	23:43
Subject Sample	# 1	0.142	23:46
Breath Volume		1.996	23:46
Air Blank		0.000	23:47
Air Blank		0.000	23:49
Subject Sample	# 2	0.139	23:50
Breath Volume		1.285	23:50
Air Blank		0.000	23:51
Control Test		0.077	23:51
Air Blank		0.000	23:52
Diagnostics Check		OK	23:52

Data Download Date/Time: 05/11/2016 11:44

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GONZALEZ,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:52

SUBJECT'S LAST NAME: CLARKE

DATE/TIME OF TEST: 04/07/2016 04:19:40

Diagnostics Check		OK	04:12
Air Blank		0.000	04:13
Control Test		0.079	04:13
Air Blank		0.000	04:14
Subject Sample	# 1	0.073	04:14
Breath Volume		2.300	04:14
Air Blank		0.000	04:15
Air Blank		0.000	04:17
Subject Sample	# 2	0.074	04:17
Breath Volume		1.652	04:17
Air Blank		0.000	04:18
Control Test		0.079	04:19
Air Blank		0.000	04:19
Diagnostics Check		OK	04:19

Data Download Date/Time: 05/11/2016 11:44

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BADEN,T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:14

SUBJECT'S LAST NAME: BORJAS ALMENDARES

DATE/TIME OF TEST: 04/09/2016 00:41:19

Diagnostics Check		OK	00:34
Air Blank		0.000	00:35
Control Test		0.079	00:35
Air Blank		0.000	00:36
Subject Sample	# 1	0.104	00:36
Breath Volume		2.082	00:36
Air Blank		0.000	00:37
Air Blank		0.000	00:39
Subject Sample	# 2	0.103	00:39
Breath Volume		2.167	00:39
Air Blank		0.000	00:40
Control Test		0.079	00:40
Air Blank		0.000	00:41
Diagnostics Check		OK	00:41

Data Download Date/Time: 05/11/2016 11:44

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SARRO,TAYLOR,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:36

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 04/09/2016 03:04:16

Diagnostics Check		OK	02:56
Air Blank		0.000	02:57
Control Test		0.079	02:57
Air Blank		0.000	02:57
Subject Sample	# 1	0.258	02:59
Breath Volume		1.476	02:59
Air Blank		0.000	02:59
Air Blank		0.000	03:01
Subject Sample	# 2	0.241	03:02
Breath Volume		1.300	03:02
Air Blank		0.000	03:03
Control Test		0.078	03:03
Air Blank		0.000	03:04
Diagnostics Check		OK	03:04

Data Download Date/Time: 05/11/2016 11:44

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY,PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRABAND,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: FLEURIMONT

DATE/TIME OF TEST: 04/09/2016 03:23:03

Diagnostics Check		OK	03:15
Air Blank		0.000	03:16
Control Test		0.078	03:16
Air Blank		0.000	03:17
Subject Sample	# 1	0.132	03:18
Breath Volume		1.535	03:18
Air Blank		0.000	03:19
Air Blank		0.000	03:21
Subject Sample	# 2	0.124	03:21
Breath Volume		1.214	03:21
Air Blank		0.000	03:22
Control Test		0.078	03:22
Air Blank		0.000	03:22
Diagnostics Check		OK	03:23

Data Download Date/Time: 05/11/2016 11:44

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BADEN,T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:35

SUBJECT'S LAST NAME: PERKINS

DATE/TIME OF TEST: 04/09/2016 05:02:12

Diagnostics Check		OK	04:55
Air Blank		0.000	04:55
Control Test		0.079	04:56
Air Blank		0.000	04:56
Subject Sample	# 1	0.158	04:57
Breath Volume		1.472	04:57
Air Blank		0.000	04:58
Air Blank		0.000	04:59
Subject Sample	# 2	0.158	05:00
Breath Volume		2.046	05:00
Air Blank		0.000	05:01
Control Test		0.078	05:01
Air Blank		0.000	05:02
Diagnostics Check		OK	05:02

Data Download Date/Time: 05/11/2016 11:44

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SEQUEIRA,N
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 19:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:59

SUBJECT'S LAST NAME: HOLLAND

DATE/TIME OF TEST: 04/09/2016 21:28:11

Diagnostics Check		OK	21:21
Air Blank		0.000	21:21
Control Test		0.080	21:22
Air Blank		0.000	21:22
Subject Sample	# 1	0.034	21:23
Breath Volume		2.234	21:23
Air Blank		0.000	21:24
Air Blank		0.000	21:25
Subject Sample	# 2	0.029	21:26
Breath Volume		1.296	21:26
Air Blank		0.000	21:27
Control Test		0.080	21:27
Air Blank		0.000	21:28
Diagnostics Check		OK	21:28

Data Download Date/Time: 05/11/2016 11:45

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK,CASEY,S
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DUNCAN,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:37

SUBJECT'S LAST NAME: GARZA

DATE/TIME OF TEST: 04/10/2016 01:04:18

Diagnostics Check		OK	00:58
Air Blank		0.000	00:58
Control Test		0.080	00:58
Air Blank		0.000	00:59
Subject Sample	# 1	0.000	00:59
Breath Volume		1.843	00:59
Air Blank		0.000	01:00
Air Blank		0.000	01:02
Subject Sample	# 2	0.000	01:02
Breath Volume		1.972	01:02
Air Blank		0.000	01:03
Control Test		0.079	01:03
Air Blank		0.000	01:04
Diagnostics Check		OK	01:04

Data Download Date/Time: 05/11/2016 11:45

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ANGELICCHI,J
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:34

SUBJECT'S LAST NAME: ESPARZA

DATE/TIME OF TEST: 04/10/2016 02:01:30

Diagnostics Check		OK	01:54
Air Blank		0.000	01:55
Control Test		0.080	01:55
Air Blank		0.000	01:56
Subject Sample	# 1	0.128	01:56
Breath Volume		3.925	01:56
Air Blank		0.000	01:57
Air Blank		0.000	01:59
Subject Sample	# 2	0.129	01:59
Breath Volume		4.164	01:59
Air Blank		0.000	02:00
Control Test		0.080	02:00
Air Blank		0.000	02:01
Diagnostics Check		OK	02:01

Data Download Date/Time: 05/11/2016 11:45

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK,CASEY,S
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: EDGE,S
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 18:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:45

SUBJECT'S LAST NAME: ANTUNEZ-SIMON

DATE/TIME OF TEST: 04/10/2016 20:20:42

Diagnostics Check		OK	20:13
Air Blank		0.000	20:14
Control Test		0.080	20:14
Air Blank		0.000	20:15
Subject Sample	# 1	0.177	20:16
Breath Volume		2.875	20:16
Air Blank		0.000	20:16
Air Blank		0.000	20:18
Subject Sample	# 2	0.166	20:19
Breath Volume		1.656	20:19
Air Blank		0.000	20:19
Control Test		0.079	20:20
Air Blank		0.000	20:20
Diagnostics Check		OK	20:20

Data Download Date/Time: 05/11/2016 11:45

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LARA,A
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:41

SUBJECT'S LAST NAME: DISHMAN

DATE/TIME OF TEST: 04/11/2016 03:09:01

Diagnostics Check		OK	03:01
Air Blank		0.000	03:02
Control Test		0.079	03:02
Air Blank		0.000	03:03
Subject Sample	# 1	0.213	03:03
Breath Volume		1.296	03:03
Air Blank		0.000	03:04
Air Blank		0.000	03:06
Subject Sample	# 2	0.212	03:07
Breath Volume		1.343	03:07
Air Blank		0.000	03:08
Control Test		0.079	03:08
Air Blank		0.000	03:08
Diagnostics Check		OK	03:09

Data Download Date/Time: 05/11/2016 11:45

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK,CASEY,S
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:07

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 04/11/2016 04:34:32

Diagnostics Check		OK	04:27
Air Blank		0.000	04:28
Control Test		0.080	04:28
Air Blank		0.000	04:29
Subject Sample	# 1	0.168	04:29
Breath Volume		2.195	04:29
Air Blank		0.000	04:30
Air Blank		0.000	04:32
Subject Sample	# 2	0.171	04:32
Breath Volume		2.183	04:32
Air Blank		0.000	04:33
Control Test		0.078	04:33
Air Blank		0.000	04:34
Diagnostics Check		OK	04:34

Data Download Date/Time: 05/11/2016 11:45

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DAWSY,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:04

SUBJECT'S LAST NAME: ALLMAN

DATE/TIME OF TEST: 04/12/2016 00:31:20

Diagnostics Check		OK	00:24
Air Blank		0.000	00:25
Control Test		0.080	00:25
Air Blank		0.000	00:26
Subject Sample	# 1	0.076	00:26
Breath Volume		2.074	00:26
Air Blank		0.000	00:27
Air Blank		0.000	00:29
Subject Sample	# 2	0.072	00:29
Breath Volume		1.308	00:29
Air Blank		0.000	00:30
Control Test		0.080	00:30
Air Blank		0.000	00:31
Diagnostics Check		OK	00:31

Data Download Date/Time: 05/11/2016 11:45

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK,CASEY,S
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:43

SUBJECT'S LAST NAME: CASTILLO

DATE/TIME OF TEST: 04/12/2016 04:13:35

Diagnostics Check		OK	04:03
Air Blank		0.000	04:04
Control Test		0.079	04:04
Air Blank		0.000	04:05
Subject Sample	# 1	0.168	04:08
Breath Volume		2.992	04:08
Air Blank		0.000	04:08
Air Blank		0.000	04:10
Subject Sample	# 2	0.159	04:11
Breath Volume		1.746	04:11
Air Blank		0.000	04:12
Control Test		0.078	04:12
Air Blank		0.000	04:13
Diagnostics Check		OK	04:13

Data Download Date/Time: 05/11/2016 11:45

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SARRO,TAYLOR,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DAWSY,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:06

SUBJECT'S LAST NAME: GARCIA-SALVADOR

DATE/TIME OF TEST: 04/13/2016 00:34:47

Diagnostics Check		OK	00:27
Air Blank		0.000	00:28
Control Test		0.079	00:28
Air Blank		0.000	00:29
Subject Sample	# 1	0.236	00:29
Breath Volume		2.183	00:29
Air Blank		0.000	00:30
Air Blank		0.000	00:32
Subject Sample	# 2	0.226	00:33
Breath Volume		2.566	00:33
Air Blank		0.000	00:33
Control Test		0.078	00:34
Air Blank		0.000	00:34
Diagnostics Check		OK	00:34

Data Download Date/Time: 05/11/2016 11:45

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY,PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GONZALEZ,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:41

SUBJECT'S LAST NAME: SMALLING

DATE/TIME OF TEST: 04/13/2016 04:09:38

Diagnostics Check		OK	04:02
Air Blank		0.000	04:03
Control Test		0.079	04:03
Air Blank		0.000	04:03
Subject Sample	# 1	0.102	04:04
Breath Volume		2.285	04:04
Air Blank		0.000	04:05
Air Blank		0.000	04:07
Subject Sample	# 2	0.101	04:07
Breath Volume		1.757	04:07
Air Blank		0.000	04:08
Control Test		0.078	04:09
Air Blank		0.000	04:09
Diagnostics Check		OK	04:09

Data Download Date/Time: 05/11/2016 11:45

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:47

SUBJECT'S LAST NAME: INGRAM

DATE/TIME OF TEST: 04/14/2016 01:14:19

Diagnostics Check		OK	01:08
Air Blank		0.000	01:08
Control Test		0.079	01:08
Air Blank		0.000	01:09
Subject Sample	# 1	0.083	01:09
Breath Volume		2.066	01:09
Air Blank		0.000	01:10
Air Blank		0.000	01:12
Subject Sample	# 2	0.076	01:12
Breath Volume		1.355	01:12
Air Blank		0.000	01:13
Control Test		0.078	01:13
Air Blank		0.000	01:14
Diagnostics Check		OK	01:14

Data Download Date/Time: 05/11/2016 11:45

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,TIMOTHY,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PALESE,A
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 14:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:22

SUBJECT'S LAST NAME: COLLIER

DATE/TIME OF TEST: 04/14/2016 16:56:34

Diagnostics Check		OK	16:42	
Air Blank		0.000	16:42	
Control Test		0.079	16:42	
Air Blank		0.000	16:43	
Subject Sample	# 1	0.084	16:46	Volume Not Met
Breath Volume		0.820	16:46	
Air Blank		0.000	16:47	
Air Blank		0.000	16:49	
Subject Sample	# 2	0.091	16:51	
Breath Volume		1.500	16:51	
Air Blank		0.000	16:52	
Air Blank		0.000	16:53	
Subject Sample	# 3	0.092	16:54	
Breath Volume		1.628	16:54	
Air Blank		0.000	16:55	
Control Test		0.078	16:55	
Air Blank		0.000	16:56	
Diagnostics Check		OK	16:56	

Data Download Date/Time: 05/11/2016 11:45

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK,CASEY,S
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DUNCAN,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:19

SUBJECT'S LAST NAME: WEBER

DATE/TIME OF TEST: 04/14/2016 23:46:30

Diagnostics Check		OK	23:39
Air Blank		0.000	23:40
Control Test		0.079	23:40
Air Blank		0.000	23:41
Subject Sample	# 1	0.192	23:41
Breath Volume		3.761	23:41
Air Blank		0.000	23:42
Air Blank		0.000	23:44
Subject Sample	# 2	0.198	23:44
Breath Volume		3.738	23:44
Air Blank		0.000	23:45
Control Test		0.078	23:45
Air Blank		0.000	23:46
Diagnostics Check		OK	23:46

Data Download Date/Time: 05/11/2016 11:45

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: THORNBERRY,S
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 17:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:36

SUBJECT'S LAST NAME: CHIJIOKE

DATE/TIME OF TEST: 04/15/2016 18:03:39

Diagnostics Check		OK	17:56
Air Blank		0.000	17:57
Control Test		0.080	17:57
Air Blank		0.000	17:58
Subject Sample	# 1	0.021	17:58
Breath Volume		2.429	17:58
Air Blank		0.000	17:59
Air Blank		0.000	18:01
Subject Sample	# 2	0.021	18:02
Breath Volume		3.164	18:02
Air Blank		0.000	18:02
Control Test		0.079	18:03
Air Blank		0.000	18:03
Diagnostics Check		OK	18:03

Data Download Date/Time: 05/11/2016 11:45

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: AGUAYO,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 16:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:55

SUBJECT'S LAST NAME: BOYETTE

DATE/TIME OF TEST: 04/15/2016 18:23:43

Diagnostics Check		OK	18:16
Air Blank		0.000	18:16
Control Test		0.080	18:16
Air Blank		0.000	18:17
Subject Sample	# 1	0.000	18:18
Breath Volume		1.328	18:18
Air Blank		0.000	18:18
Air Blank		0.000	18:20
Subject Sample	# 2	0.000	18:22
Breath Volume		1.332	18:22
Air Blank		0.000	18:22
Control Test		0.080	18:23
Air Blank		0.000	18:23
Diagnostics Check		OK	18:23

Data Download Date/Time: 05/11/2016 11:45

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FRANKLIN,R
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:39

SUBJECT'S LAST NAME: BEILHARZ

DATE/TIME OF TEST: 04/16/2016 02:06:10

Diagnostics Check		OK	01:59
Air Blank		0.000	02:00
Control Test		0.079	02:00
Air Blank		0.000	02:01
Subject Sample	# 1	0.218	02:01
Breath Volume		2.023	02:01
Air Blank		0.000	02:02
Air Blank		0.000	02:04
Subject Sample	# 2	0.220	02:04
Breath Volume		2.410	02:04
Air Blank		0.000	02:05
Control Test		0.078	02:05
Air Blank		0.000	02:06
Diagnostics Check		OK	02:06

Data Download Date/Time: 05/11/2016 11:45

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK,CASEY,S
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: CRAGG,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 05:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:30

SUBJECT'S LAST NAME: COSKUNER

DATE/TIME OF TEST: 04/16/2016 05:59:02

Diagnostics Check		OK	05:50
Air Blank		0.000	05:51
Control Test		0.079	05:51
Air Blank		0.000	05:52
Subject Sample	# 1	0.128	05:52
Breath Volume		1.199	05:52
Air Blank		0.000	05:53
Air Blank		0.000	05:55
Subject Sample	# 2	0.130	05:57
Breath Volume		1.558	05:57
Air Blank		0.000	05:58
Control Test		0.079	05:58
Air Blank		0.000	05:58
Diagnostics Check		OK	05:59

Data Download Date/Time: 05/11/2016 11:45

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMMONS,ANGELA,V
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: UNO,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 15:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:54

SUBJECT'S LAST NAME: BUSH

DATE/TIME OF TEST: 04/16/2016 16:25:00

Diagnostics Check		OK	16:14
Air Blank		0.000	16:15
Control Test		0.079	16:15
Air Blank		0.000	16:16
Subject Sample	# 1	0.247	16:18
Breath Volume		1.589	16:18
Air Blank		0.000	16:19
Air Blank		0.000	16:20
Subject Sample	# 2	0.243	16:23
Breath Volume		2.277	16:23
Air Blank		0.000	16:24
Control Test		0.077	16:24
Air Blank		0.000	16:24
Diagnostics Check		OK	16:25

Data Download Date/Time: 05/11/2016 11:45

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SARRO,TAYLOR,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VAN-TREESE,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:55

SUBJECT'S LAST NAME: GILMORE

DATE/TIME OF TEST: 04/16/2016 22:22:23

Diagnostics Check		OK	22:15
Air Blank		0.000	22:16
Control Test		0.079	22:16
Air Blank		0.000	22:17
Subject Sample	# 1	0.164	22:17
Breath Volume		1.281	22:17
Air Blank		0.000	22:18
Air Blank		0.000	22:20
Subject Sample	# 2	0.164	22:20
Breath Volume		1.363	22:20
Air Blank		0.000	22:21
Control Test		0.079	22:21
Air Blank		0.000	22:22
Diagnostics Check		OK	22:22

Data Download Date/Time: 05/11/2016 11:45

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY,PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:33

SUBJECT'S LAST NAME: SWOPE

DATE/TIME OF TEST: 04/17/2016 02:00:25

Diagnostics Check		OK	01:54
Air Blank		0.000	01:54
Control Test		0.079	01:54
Air Blank		0.000	01:55
Subject Sample	# 1	0.136	01:55
Breath Volume		1.468	01:55
Air Blank		0.000	01:56
Air Blank		0.000	01:58
Subject Sample	# 2	0.129	01:58
Breath Volume		1.515	01:58
Air Blank		0.000	01:59
Control Test		0.079	01:59
Air Blank		0.000	02:00
Diagnostics Check		OK	02:00

Data Download Date/Time: 05/11/2016 11:45

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WOODEN,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:04

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 04/17/2016 03:30:28

Diagnostics Check		OK	03:24
Air Blank		0.000	03:24
Control Test		0.079	03:25
Air Blank		0.000	03:25
Subject Sample	# 1	0.000	03:26
Breath Volume		1.839	03:26
Air Blank		0.000	03:26
Air Blank		0.000	03:28
Subject Sample	# 2	0.000	03:29
Breath Volume		2.820	03:29
Air Blank		0.000	03:29
Control Test		0.079	03:29
Air Blank		0.000	03:30
Diagnostics Check		OK	03:30

Data Download Date/Time: 05/11/2016 11:45

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY,PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNIDER,P
ARRESTING OFFICER AGENCY: Plant City PD
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:43

SUBJECT'S LAST NAME: LEWIS

DATE/TIME OF TEST: 04/17/2016 04:09:58

Diagnostics Check		OK	04:03
Air Blank		0.000	04:03
Control Test		0.080	04:04
Air Blank		0.000	04:04
Subject Sample	# 1	0.087	04:05
Breath Volume		2.734	04:05
Air Blank		0.000	04:06
Air Blank		0.000	04:07
Subject Sample	# 2	0.086	04:08
Breath Volume		3.019	04:08
Air Blank		0.000	04:09
Control Test		0.079	04:09
Air Blank		0.000	04:09
Diagnostics Check		OK	04:09

Data Download Date/Time: 05/11/2016 11:45

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FRANKLIN,R
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:06

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 04/17/2016 23:32:49

Diagnostics Check		OK	23:26
Air Blank		0.000	23:26
Control Test		0.080	23:27
Air Blank		0.000	23:27
Subject Sample	# 1	0.267	23:28
Breath Volume		1.628	23:28
Air Blank		0.000	23:29
Air Blank		0.000	23:30
Subject Sample	# 2	0.251	23:31
Breath Volume		1.308	23:31
Air Blank		0.000	23:31
Control Test		0.079	23:32
Air Blank		0.000	23:32
Diagnostics Check		OK	23:32

Data Download Date/Time: 05/11/2016 11:45

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VAN TREESE,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: BALDWIN

DATE/TIME OF TEST: 04/19/2016 03:27:14

Diagnostics Check		OK	03:20
Air Blank		0.000	03:21
Control Test		0.080	03:21
Air Blank		0.000	03:22
Subject Sample	# 1	0.067	03:22
Breath Volume		1.527	03:22
Air Blank		0.000	03:23
Air Blank		0.000	03:25
Subject Sample	# 2	0.070	03:25
Breath Volume		1.695	03:25
Air Blank		0.000	03:26
Control Test		0.079	03:26
Air Blank		0.000	03:27
Diagnostics Check		OK	03:27

Data Download Date/Time: 05/11/2016 11:45

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SARRO,TAYLOR,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARLAUG,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:19

SUBJECT'S LAST NAME: MITCHELL

DATE/TIME OF TEST: 04/19/2016 04:46:13

Diagnostics Check		OK	04:39
Air Blank		0.000	04:40
Control Test		0.079	04:40
Air Blank		0.000	04:40
Subject Sample	# 1	0.033	04:41
Breath Volume		1.152	04:41
Air Blank		0.000	04:42
Air Blank		0.000	04:44
Subject Sample	# 2	0.035	04:44
Breath Volume		1.933	04:44
Air Blank		0.000	04:45
Control Test		0.080	04:45
Air Blank		0.000	04:46
Diagnostics Check		OK	04:46

Data Download Date/Time: 05/11/2016 11:45

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: HUNTSINGER

DATE/TIME OF TEST: 04/20/2016 22:28:58

Diagnostics Check		OK	22:20
Air Blank		0.000	22:20
Control Test		0.080	22:21
Air Blank		0.000	22:21
Subject Sample	# 1	0.196	22:24
Breath Volume		1.539	22:24
Air Blank		0.000	22:25
Air Blank		0.000	22:26
Subject Sample	# 2	0.199	22:27
Breath Volume		1.679	22:27
Air Blank		0.000	22:28
Control Test		0.077	22:28
Air Blank		0.000	22:28
Diagnostics Check		OK	22:28

Data Download Date/Time: 05/11/2016 11:45

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FRANKLIN,R
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: KOSMOSKI JR

DATE/TIME OF TEST: 04/21/2016 01:47:57

Diagnostics Check		OK	01:40
Air Blank		0.000	01:41
Control Test		0.079	01:41
Air Blank		0.000	01:41
Subject Sample	# 1	0.164	01:43
Breath Volume		2.230	01:43
Air Blank		0.000	01:43
Air Blank		0.000	01:45
Subject Sample	# 2	0.164	01:46
Breath Volume		2.730	01:46
Air Blank		0.000	01:47
Control Test		0.078	01:47
Air Blank		0.000	01:47
Diagnostics Check		OK	01:47

Data Download Date/Time: 05/11/2016 11:45

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK,CASEY,S
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DAWSY,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:17

SUBJECT'S LAST NAME: RODRIGUEZ JURADO

DATE/TIME OF TEST: 04/21/2016 02:43:53

Diagnostics Check		OK	02:37
Air Blank		0.000	02:37
Control Test		0.079	02:38
Air Blank		0.000	02:38
Subject Sample	# 1	0.251	02:39
Breath Volume		1.968	02:39
Air Blank		0.000	02:40
Air Blank		0.000	02:41
Subject Sample	# 2	0.246	02:42
Breath Volume		2.347	02:42
Air Blank		0.000	02:42
Control Test		0.079	02:43
Air Blank		0.000	02:43
Diagnostics Check		OK	02:43

Data Download Date/Time: 05/11/2016 11:45

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BLANCHARD,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:43

SUBJECT'S LAST NAME: BURNS

DATE/TIME OF TEST: 04/21/2016 04:10:43

Diagnostics Check		OK	04:03
Air Blank		0.000	04:04
Control Test		0.079	04:04
Air Blank		0.000	04:05
Subject Sample	# 1	0.248	04:05
Breath Volume		2.460	04:05
Air Blank		0.000	04:06
Air Blank		0.000	04:08
Subject Sample	# 2	0.251	04:09
Breath Volume		1.515	04:09
Air Blank		0.000	04:09
Control Test		0.079	04:10
Air Blank		0.000	04:10
Diagnostics Check		OK	04:10

Data Download Date/Time: 05/11/2016 11:45

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GONZALEZ,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:14

SUBJECT'S LAST NAME: ROLDAN JR

DATE/TIME OF TEST: 04/21/2016 23:41:09

Diagnostics Check		OK	23:34
Air Blank		0.000	23:35
Control Test		0.080	23:35
Air Blank		0.000	23:35
Subject Sample	# 1	0.110	23:36
Breath Volume		2.402	23:36
Air Blank		0.000	23:37
Air Blank		0.000	23:38
Subject Sample	# 2	0.111	23:39
Breath Volume		2.953	23:39
Air Blank		0.000	23:40
Control Test		0.079	23:40
Air Blank		0.000	23:41
Diagnostics Check		OK	23:41

Data Download Date/Time: 05/11/2016 11:45

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY,PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: SALGADO

DATE/TIME OF TEST: 04/22/2016 03:42:58

Diagnostics Check		OK	03:36
Air Blank		0.000	03:36
Control Test		0.079	03:37
Air Blank		0.000	03:37
Subject Sample	# 1	0.219	03:38
Breath Volume		1.980	03:38
Air Blank		0.000	03:39
Air Blank		0.000	03:40
Subject Sample	# 2	0.216	03:41
Breath Volume		2.113	03:41
Air Blank		0.000	03:42
Control Test		0.078	03:42
Air Blank		0.000	03:42
Diagnostics Check		OK	03:42

Data Download Date/Time: 05/11/2016 11:45

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PORTMAN,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:19

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 04/23/2016 01:45:57

Diagnostics Check		OK	01:39
Air Blank		0.000	01:39
Control Test		0.080	01:40
Air Blank		0.000	01:40
Subject Sample	# 1	0.092	01:41
Breath Volume		3.152	01:41
Air Blank		0.000	01:42
Air Blank		0.000	01:43
Subject Sample	# 2	0.090	01:44
Breath Volume		2.796	01:44
Air Blank		0.000	01:45
Control Test		0.079	01:45
Air Blank		0.000	01:45
Diagnostics Check		OK	01:45

Data Download Date/Time: 05/11/2016 11:45

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:47

SUBJECT'S LAST NAME: JONES JR

DATE/TIME OF TEST: 04/23/2016 02:13:38

Diagnostics Check		OK	02:07
Air Blank		0.000	02:07
Control Test		0.080	02:07
Air Blank		0.000	02:08
Subject Sample	# 1	0.174	02:09
Breath Volume		1.753	02:09
Air Blank		0.000	02:09
Air Blank		0.000	02:11
Subject Sample	# 2	0.171	02:11
Breath Volume		1.515	02:11
Air Blank		0.000	02:12
Control Test		0.079	02:13
Air Blank		0.000	02:13
Diagnostics Check		OK	02:13

Data Download Date/Time: 05/11/2016 11:45

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY,PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PORTMAN,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:14

SUBJECT'S LAST NAME: EISENZIMMER

DATE/TIME OF TEST: 04/23/2016 03:41:01

Diagnostics Check		OK	03:34
Air Blank		0.000	03:34
Control Test		0.080	03:35
Air Blank		0.000	03:35
Subject Sample	# 1	0.094	03:36
Breath Volume		2.562	03:36
Air Blank		0.000	03:37
Air Blank		0.000	03:38
Subject Sample	# 2	0.093	03:39
Breath Volume		3.246	03:39
Air Blank		0.000	03:40
Control Test		0.079	03:40
Air Blank		0.000	03:40
Diagnostics Check		OK	03:41

Data Download Date/Time: 05/11/2016 11:45

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK,CASEY,S
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:18

SUBJECT'S LAST NAME: MAUK

DATE/TIME OF TEST: 04/23/2016 22:46:22

Diagnostics Check		OK	22:39
Air Blank		0.000	22:40
Control Test		0.079	22:40
Air Blank		0.000	22:41
Subject Sample	# 1	0.098	22:41
Breath Volume		1.300	22:41
Air Blank		0.000	22:42
Air Blank		0.000	22:44
Subject Sample	# 2	0.106	22:44
Breath Volume		1.347	22:44
Air Blank		0.000	22:45
Control Test		0.078	22:45
Air Blank		0.000	22:46
Diagnostics Check		OK	22:46

Data Download Date/Time: 05/11/2016 11:45

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GONZALEZ,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:39

SUBJECT'S LAST NAME: KELLY

DATE/TIME OF TEST: 04/24/2016 02:07:29

Diagnostics Check		OK	02:00
Air Blank		0.000	02:01
Control Test		0.079	02:01
Air Blank		0.000	02:01
Subject Sample	# 1	0.256	02:02
Breath Volume		1.882	02:02
Air Blank		0.000	02:03
Air Blank		0.000	02:05
Subject Sample	# 2	0.255	02:05
Breath Volume		1.347	02:05
Air Blank		0.000	02:06
Control Test		0.078	02:06
Air Blank		0.000	02:07
Diagnostics Check		OK	02:07

Data Download Date/Time: 05/11/2016 11:45

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK,CASEY,S
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VAN-TREESE,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: IRIZARRY GARCIA

DATE/TIME OF TEST: 04/24/2016 03:02:40

Diagnostics Check		OK	02:56
Air Blank		0.000	02:56
Control Test		0.078	02:57
Air Blank		0.000	02:57
Subject Sample	# 1	0.084	02:58
Breath Volume		2.242	02:58
Air Blank		0.000	02:58
Air Blank		0.000	03:00
Subject Sample	# 2	0.081	03:01
Breath Volume		1.429	03:01
Air Blank		0.000	03:01
Control Test		0.079	03:02
Air Blank		0.000	03:02
Diagnostics Check		OK	03:02

Data Download Date/Time: 05/11/2016 11:45

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ROSE,C
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:41

SUBJECT'S LAST NAME: CRUZ

DATE/TIME OF TEST: 04/24/2016 04:12:46

Diagnostics Check		OK	04:02
Air Blank		0.000	04:02
Control Test		0.079	04:02
Air Blank		0.000	04:03
Subject Sample	# 1	0.225	04:04
Breath Volume		3.726	04:04
Air Blank		0.000	04:04
Air Blank		0.000	04:06
Subject Sample	# 2	0.017	04:07
Breath Volume		1.742	04:07
Air Blank		0.000	04:08
Air Blank		0.000	04:10
Subject Sample	# 3	0.207	04:11
Breath Volume		3.636	04:11
Air Blank		0.000	04:11
Control Test		0.078	04:12
Air Blank		0.000	04:12
Diagnostics Check		OK	04:12

Data Download Date/Time: 05/11/2016 11:45

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DUNCAN,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:44

SUBJECT'S LAST NAME: SALVADOR-TEMOXTLE

DATE/TIME OF TEST: 04/24/2016 05:13:53

Diagnostics Check		OK	05:04
Air Blank		0.000	05:05
Control Test		0.078	05:05
Air Blank		0.000	05:06
Subject Sample	# 1	0.188	05:07
Breath Volume		1.937	05:07
Air Blank		0.000	05:08
Air Blank		0.000	05:10
Subject Sample	# 2	0.182	05:12
Breath Volume		2.535	05:12
Air Blank		0.000	05:12
Control Test		0.078	05:13
Air Blank		0.000	05:13
Diagnostics Check		OK	05:13

Data Download Date/Time: 05/11/2016 11:45

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK,CASEY,S
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: BARRA

DATE/TIME OF TEST: 04/25/2016 04:13:42

Diagnostics Check		OK	04:05
Air Blank		0.000	04:06
Control Test		0.079	04:06
Air Blank		0.000	04:07
Subject Sample	# 1	0.108	04:08
Breath Volume		1.164	04:08
Air Blank		0.000	04:09
Air Blank		0.000	04:11
Subject Sample	# 2	0.108	04:12
Breath Volume		2.035	04:12
Air Blank		0.000	04:12
Control Test		0.078	04:13
Air Blank		0.000	04:13
Diagnostics Check		OK	04:13

Data Download Date/Time: 05/11/2016 11:45

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: CRESCENZI,T
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:26

SUBJECT'S LAST NAME: OZGOL

DATE/TIME OF TEST: 04/26/2016 00:53:18

Diagnostics Check		OK	00:46
Air Blank		0.000	00:47
Control Test		0.080	00:47
Air Blank		0.000	00:47
Subject Sample	# 1	0.231	00:48
Breath Volume		2.140	00:48
Air Blank		0.000	00:49
Air Blank		0.000	00:51
Subject Sample	# 2	0.233	00:51
Breath Volume		2.117	00:51
Air Blank		0.000	00:52
Control Test		0.079	00:52
Air Blank		0.000	00:53
Diagnostics Check		OK	00:53

Data Download Date/Time: 05/11/2016 11:45

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:19

SUBJECT'S LAST NAME: DENSON

DATE/TIME OF TEST: 04/26/2016 04:48:27

Diagnostics Check		OK	04:39
Air Blank		0.000	04:40
Control Test		0.079	04:40
Air Blank		0.000	04:41
Subject Sample	# 1	0.042	04:43
Breath Volume		1.132	04:43
Air Blank		0.000	04:43
Air Blank		0.000	04:45
Subject Sample	# 2	0.042	04:46
Breath Volume		1.496	04:46
Air Blank		0.000	04:47
Control Test		0.080	04:47
Air Blank		0.000	04:48
Diagnostics Check		OK	04:48

Data Download Date/Time: 05/11/2016 11:45

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARLAUG,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: FERNANDEZ

DATE/TIME OF TEST: 04/27/2016 02:22:04

Diagnostics Check		OK	02:15
Air Blank		0.000	02:16
Control Test		0.080	02:16
Air Blank		0.000	02:17
Subject Sample	# 1	0.079	02:17
Breath Volume		5.082	02:17
Air Blank		0.000	02:18
Air Blank		0.000	02:20
Subject Sample	# 2	0.079	02:20
Breath Volume		4.796	02:20
Air Blank		0.000	02:21
Control Test		0.080	02:21
Air Blank		0.000	02:21
Diagnostics Check		OK	02:22

Data Download Date/Time: 05/11/2016 11:45

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SARRO,TAYLOR,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DUNCAN,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:42

SUBJECT'S LAST NAME: MARRIOTT

DATE/TIME OF TEST: 04/27/2016 21:10:09

Diagnostics Check		OK	21:02
Air Blank		0.000	21:03
Control Test		0.080	21:03
Air Blank		0.000	21:04
Subject Sample	# 1	0.145	21:05
Breath Volume		1.972	21:05
Air Blank		0.000	21:06
Air Blank		0.000	21:07
Subject Sample	# 2	0.145	21:08
Breath Volume		1.207	21:08
Air Blank		0.000	21:09
Control Test		0.080	21:09
Air Blank		0.000	21:10
Diagnostics Check		OK	21:10

Data Download Date/Time: 05/11/2016 11:45

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PALFY,R
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:19

SUBJECT'S LAST NAME: HERNANDEZ UBALDO

DATE/TIME OF TEST: 04/28/2016 02:46:41

Diagnostics Check		OK	02:39
Air Blank		0.000	02:39
Control Test		0.080	02:39
Air Blank		0.000	02:40
Subject Sample	# 1	0.264	02:41
Breath Volume		2.449	02:41
Air Blank		0.000	02:41
Air Blank		0.000	02:43
Subject Sample	# 2	0.255	02:44
Breath Volume		2.007	02:44
Air Blank		0.000	02:45
Control Test		0.079	02:46
Air Blank		0.000	02:46
Diagnostics Check		OK	02:46

Data Download Date/Time: 05/11/2016 11:45

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY,PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SIMPKINS,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:42

SUBJECT'S LAST NAME: CHAMBERS

DATE/TIME OF TEST: 04/28/2016 05:09:10

Diagnostics Check		OK	05:02
Air Blank		0.000	05:02
Control Test		0.080	05:02
Air Blank		0.000	05:03
Subject Sample	# 1	0.136	05:04
Breath Volume		1.824	05:04
Air Blank		0.000	05:05
Air Blank		0.000	05:06
Subject Sample	# 2	0.131	05:07
Breath Volume		1.800	05:07
Air Blank		0.000	05:08
Control Test		0.079	05:08
Air Blank		0.000	05:09
Diagnostics Check		OK	05:09

Data Download Date/Time: 05/11/2016 11:45

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK,CASEY,S
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: AGUAYO,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:38

SUBJECT'S LAST NAME: GRIST

DATE/TIME OF TEST: 04/29/2016 01:06:21

Diagnostics Check		OK	00:59
Air Blank		0.000	01:00
Control Test		0.079	01:00
Air Blank		0.000	01:00
Subject Sample	# 1	0.146	01:01
Breath Volume		1.292	01:01
Air Blank		0.000	01:02
Air Blank		0.000	01:03
Subject Sample	# 2	0.151	01:04
Breath Volume		2.339	01:04
Air Blank		0.000	01:05
Control Test		0.078	01:05
Air Blank		0.000	01:06
Diagnostics Check		OK	01:06

Data Download Date/Time: 05/11/2016 11:45

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:37

SUBJECT'S LAST NAME: ZEWALK

DATE/TIME OF TEST: 04/29/2016 02:04:59

Diagnostics Check		OK	01:57
Air Blank		0.000	01:58
Control Test		0.079	01:58
Air Blank		0.000	01:59
Subject Sample	# 1	0.151	02:00
Breath Volume		1.289	02:00
Air Blank		0.000	02:00
Air Blank		0.000	02:02
Subject Sample	# 2	0.160	02:03
Breath Volume		1.917	02:03
Air Blank		0.000	02:04
Control Test		0.078	02:04
Air Blank		0.000	02:04
Diagnostics Check		OK	02:04

Data Download Date/Time: 05/11/2016 11:45

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BLANCHARD,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:26

SUBJECT'S LAST NAME: SOSA

DATE/TIME OF TEST: 04/29/2016 23:54:07

Diagnostics Check		OK	23:46
Air Blank		0.000	23:47
Control Test		0.080	23:47
Air Blank		0.000	23:47
Subject Sample	# 1	0.128	23:49
Breath Volume		1.253	23:49
Air Blank		0.000	23:50
Air Blank		0.000	23:51
Subject Sample	# 2	0.138	23:52
Breath Volume		1.953	23:52
Air Blank		0.000	23:53
Control Test		0.078	23:53
Air Blank		0.000	23:54
Diagnostics Check		OK	23:54

Data Download Date/Time: 05/11/2016 11:45

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK,CASEY,S
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ANGELICCHI,J
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:07

SUBJECT'S LAST NAME: CARSON

DATE/TIME OF TEST: 04/30/2016 02:34:06

Diagnostics Check		OK	02:27
Air Blank		0.000	02:28
Control Test		0.080	02:28
Air Blank		0.000	02:29
Subject Sample	# 1	0.188	02:29
Breath Volume		3.273	02:29
Air Blank		0.000	02:30
Air Blank		0.000	02:31
Subject Sample	# 2	0.192	02:32
Breath Volume		3.144	02:32
Air Blank		0.000	02:33
Control Test		0.081	02:33
Air Blank		0.000	02:34
Diagnostics Check		OK	02:34

Data Download Date/Time: 05/11/2016 11:45

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK,CASEY,S
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PORTMAN,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:52

SUBJECT'S LAST NAME: PHELPS

DATE/TIME OF TEST: 04/30/2016 03:20:38

Diagnostics Check		OK	03:14
Air Blank		0.000	03:14
Control Test		0.080	03:15
Air Blank		0.000	03:15
Subject Sample	# 1	0.099	03:16
Breath Volume		1.289	03:16
Air Blank		0.000	03:16
Air Blank		0.000	03:18
Subject Sample	# 2	0.096	03:19
Breath Volume		1.605	03:19
Air Blank		0.000	03:19
Control Test		0.079	03:20
Air Blank		0.000	03:20
Diagnostics Check		OK	03:20

Data Download Date/Time: 05/11/2016 11:45

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY,PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FRANKLIN,R
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 19:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:41

SUBJECT'S LAST NAME: RADTKE

DATE/TIME OF TEST: 04/30/2016 21:17:24

Diagnostics Check		OK	21:01	
Air Blank		0.000	21:01	
Control Test		0.079	21:02	
Air Blank		0.000	21:02	
Subject Sample	# 1	0.186	21:06	
Breath Volume		1.398	21:06	
Air Blank		0.000	21:06	
Air Blank		0.000	21:08	
Subject Sample	# 2	0.157	21:11	Volume Not Met
Breath Volume		0.675	21:11	
Air Blank		0.000	21:12	
Air Blank		0.000	21:14	
Subject Sample	# 3	0.183	21:15	
Breath Volume		1.476	21:15	
Air Blank		0.000	21:16	
Control Test		0.076	21:16	
Air Blank		0.000	21:17	
Diagnostics Check		OK	21:17	

Data Download Date/Time: 05/11/2016 11:45

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: AGUAYO,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: FONT JR

DATE/TIME OF TEST: 04/30/2016 23:15:25

Diagnostics Check		OK	23:09
Air Blank		0.000	23:09
Control Test		0.078	23:09
Air Blank		0.000	23:10
Subject Sample	# 1	0.264	23:10
Breath Volume		3.074	23:10
Air Blank		0.000	23:11
Air Blank		0.000	23:13
Subject Sample	# 2	0.265	23:13
Breath Volume		2.949	23:13
Air Blank		0.000	23:14
Control Test		0.078	23:14
Air Blank		0.000	23:15
Diagnostics Check		OK	23:15

Data Download Date/Time: 05/11/2016 11:45

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEVENS,MITCHELL,M
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: STEVENS,MITCHELL
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 03/27/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:22

SUBJECT'S LAST NAME: SONSALLA

DATE/TIME OF TEST: 04/01/2016 04:50:58

Diagnostics Check		OK	04:42	
Air Blank		0.000	04:42	
Control Test		0.081	04:43	
Air Blank		0.000	04:43	
Subject Sample	# 1		04:46	No Sample Provided
Breath Volume		0.000	04:46	
Air Blank		0.000	04:47	
Air Blank		0.000	04:49	
Subject Sample	# 2		04:49	Subject Test Refused
Breath Volume		0.000	04:49	
Air Blank		0.000	04:50	
Control Test		0.081	04:50	
Air Blank		0.000	04:50	
Diagnostics Check		OK	04:50	

Data Download Date/Time: 04/13/2016 13:31

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEEKS,JEREMIAH,S
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: MEEKS,JEREMIAH
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 03/27/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:28

SUBJECT'S LAST NAME: GORMLEY

DATE/TIME OF TEST: 04/03/2016 01:54:52

Diagnostics Check		OK	01:48
Air Blank		0.000	01:48
Control Test		0.080	01:48
Air Blank		0.000	01:49
Subject Sample	# 1	0.103	01:49
Breath Volume		3.835	01:49
Air Blank		0.000	01:50
Air Blank		0.000	01:52
Subject Sample	# 2	0.109	01:53
Breath Volume		3.339	01:53
Air Blank		0.000	01:53
Control Test		0.079	01:54
Air Blank		0.000	01:54
Diagnostics Check		OK	01:54

Data Download Date/Time: 04/13/2016 13:31

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHELLEY,SAMUEL,L
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: SHELLEY,SAMUEL
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 03/27/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: ADAMS

DATE/TIME OF TEST: 04/03/2016 03:46:59

Diagnostics Check		OK	03:40
Air Blank		0.000	03:40
Control Test		0.081	03:41
Air Blank		0.000	03:41
Subject Sample	# 1	0.092	03:42
Breath Volume		3.398	03:42
Air Blank		0.000	03:43
Air Blank		0.000	03:44
Subject Sample	# 2	0.088	03:45
Breath Volume		2.566	03:45
Air Blank		0.000	03:46
Control Test		0.081	03:46
Air Blank		0.000	03:46
Diagnostics Check		OK	03:46

Data Download Date/Time: 04/13/2016 13:31

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DURHAM,JAMES,S
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: DURHAM,JAMES
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 03/27/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: JOYCE

DATE/TIME OF TEST: 04/05/2016 23:39:53

Diagnostics Check		OK	23:32
Air Blank		0.000	23:32
Control Test		0.081	23:32
Air Blank		0.000	23:33
Subject Sample	# 1	0.000	23:35
Breath Volume		1.824	23:35
Air Blank		0.000	23:35
Air Blank		0.000	23:37
Subject Sample	# 2	0.000	23:38
Breath Volume		1.574	23:38
Air Blank		0.000	23:38
Control Test		0.080	23:39
Air Blank		0.000	23:39
Diagnostics Check		OK	23:39

Data Download Date/Time: 04/13/2016 13:31

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEETERS,OSBORNE,H
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: TEETERS,OSBORNE
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 03/27/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: WEAVER

DATE/TIME OF TEST: 04/10/2016 01:04:06

Diagnostics Check		OK	00:59	
Air Blank		0.000	00:59	
Control Test		0.082	00:59	
Air Blank		0.000	01:00	
Subject Sample	# 1		01:02	Subject Test Refused
Breath Volume		0.000	01:02	
Air Blank		0.000	01:03	
Control Test		0.082	01:03	
Air Blank		0.000	01:04	
Diagnostics Check		OK	01:04	

Data Download Date/Time: 04/13/2016 13:31

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEETERS,OSBORNE,H
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: TEETERS,OSBORNE
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 03/27/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: WEAVER

DATE/TIME OF TEST: 04/10/2016 01:12:29

Diagnostics Check		OK	01:05
Air Blank		0.000	01:06
Control Test		0.081	01:06
Air Blank		0.000	01:07
Subject Sample	# 1	0.100	01:07
Breath Volume		1.722	01:07
Air Blank		0.000	01:08
Air Blank		0.000	01:10
Subject Sample	# 2	0.102	01:10
Breath Volume		1.765	01:10
Air Blank		0.000	01:11
Control Test		0.082	01:11
Air Blank		0.000	01:12
Diagnostics Check		OK	01:12

Data Download Date/Time: 04/13/2016 13:31

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUDSON,JOSHUA,K
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: GARY,ADAM
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 04/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: MITCHELL

DATE/TIME OF TEST: 04/28/2016 02:31:31

Diagnostics Check		OK	02:24
Air Blank		0.000	02:25
Control Test		0.080	02:25
Air Blank		0.000	02:26
Subject Sample	# 1	0.000	02:27
Breath Volume		3.281	02:27
Air Blank		0.000	02:27
Air Blank		0.000	02:29
Subject Sample	# 2	0.000	02:30
Breath Volume		1.550	02:30
Air Blank		0.000	02:30
Control Test		0.081	02:30
Air Blank		0.000	02:31
Diagnostics Check		OK	02:31

Data Download Date/Time: 05/21/2016 04:52

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUDSON,JOSHUA,K
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: HUDSON,JOSHUA
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 04/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:46

SUBJECT'S LAST NAME: GONZALES

DATE/TIME OF TEST: 04/30/2016 02:18:13

Diagnostics Check		OK	02:15	
Air Blank		0.000	02:15	
Control Test		0.081	02:15	
Air Blank		0.000	02:16	
Subject Sample	# 1		02:16	Subject Test Refused
Breath Volume		0.000	02:16	
Air Blank		0.000	02:17	
Control Test		0.081	02:17	
Air Blank		0.000	02:18	
Diagnostics Check		OK	02:18	

Data Download Date/Time: 05/21/2016 04:52

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VILLAFANA,CESAR,J
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: VILAFANA,CESAR
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 03/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: BECERRA

DATE/TIME OF TEST: 04/01/2016 00:56:23

Diagnostics Check		OK	00:53	
Air Blank		0.000	00:53	
Control Test		0.078	00:54	
Air Blank		0.000	00:54	
Subject Sample	# 1		00:54	Subject Test Refused
Breath Volume		0.000	00:54	
Air Blank		0.000	00:55	
Control Test		0.077	00:55	
Air Blank		0.000	00:56	
Diagnostics Check		OK	00:56	

Data Download Date/Time: 04/22/2016 18:24

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHERRELUS,RUDDENS,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: CHERRELUS,RUDDENS
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 03/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 05:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:27

SUBJECT'S LAST NAME: KELLY

DATE/TIME OF TEST: 04/01/2016 05:54:38

Diagnostics Check		OK	05:48
Air Blank		0.000	05:48
Control Test		0.079	05:48
Air Blank		0.000	05:49
Subject Sample	# 1	0.228	05:50
Breath Volume		2.195	05:50
Air Blank		0.000	05:50
Air Blank		0.000	05:52
Subject Sample	# 2	0.230	05:53
Breath Volume		1.453	05:53
Air Blank		0.000	05:53
Control Test		0.079	05:54
Air Blank		0.000	05:54
Diagnostics Check		OK	05:54

Data Download Date/Time: 04/22/2016 18:24

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRADDY,STEVEN,A
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: BRADDY,STEVEN
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 03/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 07:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:58

SUBJECT'S LAST NAME: GUEVARA

DATE/TIME OF TEST: 04/02/2016 08:21:12

Diagnostics Check		OK	08:18	
Air Blank		0.000	08:18	
Control Test		0.078	08:19	
Air Blank		0.000	08:19	
Subject Sample	# 1		08:19	Subject Test Refused
Breath Volume		0.000	08:19	
Air Blank		0.000	08:20	
Control Test		0.079	08:20	
Air Blank		0.000	08:21	
Diagnostics Check		OK	08:21	

Data Download Date/Time: 04/22/2016 18:24

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CABEZA,HIRAM,E
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: CABEZA,HIRAM
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 03/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 19:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:40

SUBJECT'S LAST NAME: GUTIERREZ TALAVERA

DATE/TIME OF TEST: 04/02/2016 20:06:51

Diagnostics Check		OK	20:00
Air Blank		0.000	20:01
Control Test		0.079	20:01
Air Blank		0.000	20:02
Subject Sample	# 1	0.185	20:02
Breath Volume		2.113	20:02
Air Blank		0.000	20:03
Air Blank		0.000	20:04
Subject Sample	# 2	0.181	20:05
Breath Volume		1.941	20:05
Air Blank		0.000	20:05
Control Test		0.079	20:06
Air Blank		0.000	20:06
Diagnostics Check		OK	20:06

Data Download Date/Time: 04/22/2016 18:24

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHERRELUS,RUDDENS,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: CHERRELUS,RUDDENS
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 03/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:56

SUBJECT'S LAST NAME: MARSHALL

DATE/TIME OF TEST: 04/03/2016 02:22:39

Diagnostics Check		OK	02:16
Air Blank		0.000	02:16
Control Test		0.079	02:17
Air Blank		0.000	02:17
Subject Sample	# 1	0.210	02:18
Breath Volume		1.300	02:18
Air Blank		0.000	02:18
Air Blank		0.000	02:20
Subject Sample	# 2	0.210	02:21
Breath Volume		1.242	02:21
Air Blank		0.000	02:21
Control Test		0.079	02:22
Air Blank		0.000	02:22
Diagnostics Check		OK	02:22

Data Download Date/Time: 04/22/2016 18:24

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TADLOCK,JAYNIS,L
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: TADLOCK,J
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 03/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: HERRERA JR

DATE/TIME OF TEST: 04/03/2016 03:52:55

Diagnostics Check	OK	03:50	
Air Blank	0.000	03:51	
Control Test	0.077	03:51	
Air Blank		03:52	Purge Fail
Air Blank	0.000	03:52	

Data Download Date/Time: 04/22/2016 18:24

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TADLOCK,J,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: TADLOCK,J
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 03/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: HERRERA JR

DATE/TIME OF TEST: 04/03/2016 04:12:01

Diagnostics Check		OK	03:54	
Air Blank		0.000	03:55	
Control Test		0.079	03:55	
Air Blank		0.000	03:56	
Subject Sample	# 1	0.231	03:59	Volume Not Met
Breath Volume		0.640	03:59	
Air Blank		0.000	04:00	
Air Blank		0.000	04:01	
Subject Sample	# 2	0.225	04:04	
Breath Volume		1.796	04:04	
Air Blank		0.000	04:05	
Air Blank		0.000	04:07	
Subject Sample	# 3	0.212	04:10	
Breath Volume		1.371	04:10	
Air Blank		0.000	04:11	
Control Test		0.073	04:11	Control Outside Tolerance
Air Blank		0.000	04:12	

Data Download Date/Time: 04/22/2016 18:24

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOMINGO,ROLANDO,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: DOMINGO,ROLANDO
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 03/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: FALLAS

DATE/TIME OF TEST: 04/03/2016 04:27:54

Diagnostics Check		OK	04:21
Air Blank		0.000	04:22
Control Test		0.077	04:22
Air Blank		0.000	04:23
Subject Sample	# 1	0.207	04:23
Breath Volume		1.953	04:23
Air Blank		0.000	04:24
Air Blank		0.000	04:26
Subject Sample	# 2	0.207	04:26
Breath Volume		1.824	04:26
Air Blank		0.000	04:26
Control Test		0.078	04:27
Air Blank		0.000	04:27
Diagnostics Check		OK	04:27

Data Download Date/Time: 04/22/2016 18:24

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TADLOCK,J,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: TADLOCK,J
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 03/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: HERRERA JR

DATE/TIME OF TEST: 04/03/2016 05:03:57

Diagnostics Check		OK	04:51	
Air Blank		0.000	04:51	
Control Test		0.078	04:51	
Air Blank		0.000	04:52	
Subject Sample	# 1	0.207	04:53	
Breath Volume		1.281	04:53	
Air Blank		0.000	04:54	
Air Blank		0.000	04:56	
Subject Sample	# 2	0.209	04:59	Volume Not Met
Breath Volume		0.437	04:59	
Air Blank		0.000	05:00	
Air Blank		0.000	05:01	
Subject Sample	# 3	0.215	05:02	
Breath Volume		1.371	05:02	
Air Blank		0.000	05:03	
Control Test		0.072	05:03	Control Outside Tolerance
Air Blank		0.000	05:03	

Data Download Date/Time: 04/22/2016 18:24

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TADLOCK,J,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: TADLOCK,J
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 03/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: HERRERA JR

DATE/TIME OF TEST: 04/03/2016 05:11:36

Diagnostics Check		OK	05:07	
Air Blank		0.000	05:07	
Control Test		0.075	05:08	
Air Blank		0.000	05:08	
Subject Sample	# 1		05:09	Subject Test Refused
Breath Volume		0.140	05:09	
Air Blank		0.000	05:10	
Control Test		0.077	05:10	
Air Blank		0.000	05:11	
Diagnostics Check		OK	05:11	

Data Download Date/Time: 04/22/2016 18:24

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORTIZ,LUIS,A
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: MENDEZ,ERNESTO
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 03/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 04/05/2016 21:50:55

Diagnostics Check		OK	21:41	
Air Blank		0.000	21:42	
Control Test		0.078	21:42	
Air Blank		0.000	21:43	
Subject Sample	# 1		21:46	No Sample Provided
Breath Volume		0.000	21:46	
Air Blank		0.000	21:46	
Air Blank		0.000	21:48	
Subject Sample	# 2		21:49	Subject Test Refused
Breath Volume		0.000	21:49	
Air Blank		0.000	21:49	
Control Test		0.079	21:50	
Air Blank		0.000	21:50	
Diagnostics Check		OK	21:50	

Data Download Date/Time: 04/22/2016 18:24

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MATHIS,ANDRE,D
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: MATHIS,ANDRE
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 03/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:23

SUBJECT'S LAST NAME: ALONSO PEREZ

DATE/TIME OF TEST: 04/08/2016 21:50:47

Diagnostics Check	OK	21:49	
Air Blank	0.000	21:49	
Control Test	0.078	21:50	RFI Detect
Air Blank	0.000	21:50	

Data Download Date/Time: 04/22/2016 18:24

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MATHIS,ANDRE,D
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: MATHIS,ANDRE
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 03/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:23

SUBJECT'S LAST NAME: ALONSO PEREZ

DATE/TIME OF TEST: 04/08/2016 21:57:49

Diagnostics Check		OK	21:53	
Air Blank		0.000	21:54	
Control Test		0.078	21:54	
Air Blank		0.000	21:55	
Subject Sample	# 1		21:56	Subject Test Refused
Breath Volume		0.000	21:56	
Air Blank		0.000	21:56	
Control Test		0.079	21:57	
Air Blank		0.000	21:57	
Diagnostics Check		OK	21:57	

Data Download Date/Time: 04/22/2016 18:24

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOMINGO,ROLANDO,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: PHILIPPE,W
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 03/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: ALVARADO

DATE/TIME OF TEST: 04/09/2016 00:59:58

Diagnostics Check		OK	00:52
Air Blank		0.000	00:53
Control Test		0.078	00:53
Air Blank		0.000	00:54
Subject Sample	# 1	0.143	00:55
Breath Volume		1.199	00:55
Air Blank		0.000	00:55
Air Blank		0.000	00:57
Subject Sample	# 2	0.142	00:58
Breath Volume		1.664	00:58
Air Blank		0.000	00:59
Control Test		0.078	00:59
Air Blank		0.000	00:59
Diagnostics Check		OK	00:59

Data Download Date/Time: 04/22/2016 18:24

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHERRELUS,RUDDENS,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: ANTELO,ALEXANDER
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 03/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: GRIMM

DATE/TIME OF TEST: 04/09/2016 03:14:56

Diagnostics Check		OK	03:11	
Air Blank		0.000	03:12	
Control Test		0.078	03:12	
Air Blank		0.000	03:13	
Subject Sample	# 1		03:13	Subject Test Refused
Breath Volume		0.000	03:13	
Air Blank		0.000	03:14	
Control Test		0.078	03:14	
Air Blank		0.000	03:14	
Diagnostics Check		OK	03:14	

Data Download Date/Time: 04/22/2016 18:24

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUZMAN,JONATHAN,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: GUZMAN,JONATHAN
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 03/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 08:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:05

SUBJECT'S LAST NAME: CANALES

DATE/TIME OF TEST: 04/09/2016 09:42:44

Diagnostics Check		OK	09:32
Air Blank		0.000	09:33
Control Test		0.078	09:33
Air Blank		0.000	09:34
Subject Sample	# 1	0.215	09:35
Breath Volume		1.562	09:35
Air Blank		0.000	09:35
Air Blank		0.000	09:37
Subject Sample	# 2	0.240	09:38
Breath Volume		2.812	09:38
Air Blank		0.000	09:38
Air Blank		0.000	09:40
Subject Sample	# 3	0.228	09:41
Breath Volume		3.781	09:41
Air Blank		0.000	09:41
Control Test		0.078	09:42
Air Blank		0.000	09:42
Diagnostics Check		OK	09:42

Data Download Date/Time: 04/22/2016 18:24

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LLOVET,GERARDO,A
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: PAIGO,RYAN
ARRESTING OFFICER AGENCY: Key Biscayne PD
LAST AGENCY INSPECTION: 03/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:49

SUBJECT'S LAST NAME: BASSI

DATE/TIME OF TEST: 04/10/2016 00:13:17

Diagnostics Check		OK	00:10	
Air Blank		0.000	00:10	
Control Test		0.079	00:11	
Air Blank		0.000	00:11	
Subject Sample	# 1		00:11	Subject Test Refused
Breath Volume		0.000	00:11	
Air Blank		0.000	00:12	
Control Test		0.078	00:12	
Air Blank		0.000	00:13	
Diagnostics Check		OK	00:13	

Data Download Date/Time: 04/22/2016 18:24

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOMINGO,ROLANDO,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: NODAR,A
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 03/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: LOPEZ GARCIA

DATE/TIME OF TEST: 04/10/2016 01:25:19

Diagnostics Check		OK	01:22	
Air Blank		0.000	01:22	
Control Test		0.078	01:23	
Air Blank		0.000	01:23	
Subject Sample	# 1		01:23	Subject Test Refused
Breath Volume		0.000	01:23	
Air Blank		0.000	01:24	
Control Test		0.078	01:24	
Air Blank		0.000	01:25	
Diagnostics Check		OK	01:25	

Data Download Date/Time: 04/22/2016 18:24

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHERRELUS,RUDDENS,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: CHERRELUS,RUDDENS
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 03/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: ALHOWAISH

DATE/TIME OF TEST: 04/10/2016 03:15:50

Diagnostics Check		OK	03:06
Air Blank		0.000	03:06
Control Test		0.078	03:07
Air Blank		0.000	03:07
Subject Sample	# 1	0.165	03:10
Breath Volume		1.593	03:10
Air Blank		0.000	03:10
Air Blank		0.000	03:12
Subject Sample	# 2	0.158	03:14
Breath Volume		1.589	03:14
Air Blank		0.000	03:14
Control Test		0.078	03:15
Air Blank		0.000	03:15
Diagnostics Check		OK	03:15

Data Download Date/Time: 04/22/2016 18:24

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,JASIEL,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: TARGIA,FRANK
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 03/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:11

SUBJECT'S LAST NAME: MOJICA

DATE/TIME OF TEST: 04/11/2016 03:38:31

Diagnostics Check		OK	03:31
Air Blank		0.000	03:31
Control Test		0.078	03:32
Air Blank		0.000	03:32
Subject Sample	# 1	0.079	03:33
Breath Volume		1.898	03:33
Air Blank		0.000	03:34
Air Blank		0.000	03:36
Subject Sample	# 2	0.079	03:36
Breath Volume		2.136	03:36
Air Blank		0.000	03:37
Control Test		0.078	03:37
Air Blank		0.000	03:38
Diagnostics Check		OK	03:38

Data Download Date/Time: 04/22/2016 18:24

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HEREDIA RUBIO,ROBERTO,A
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: HEREDIA RUBIO,ROBERTO
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 03/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:09

SUBJECT'S LAST NAME: JIMENEZ

DATE/TIME OF TEST: 04/11/2016 22:34:04

Diagnostics Check		OK	22:30	
Air Blank		0.000	22:31	
Control Test		0.077	22:31	
Air Blank		0.000	22:31	
Subject Sample	# 1		22:32	Subject Test Refused
Breath Volume		0.000	22:32	
Air Blank		0.000	22:33	
Control Test		0.078	22:33	
Air Blank		0.000	22:33	
Diagnostics Check		OK	22:34	

Data Download Date/Time: 04/22/2016 18:25

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,LUIS,G
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: RODRIGUEZ,LUIS
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 03/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:07

SUBJECT'S LAST NAME: SCOLA

DATE/TIME OF TEST: 04/12/2016 01:40:28

Diagnostics Check		OK	01:36	
Air Blank		0.000	01:36	
Control Test		0.078	01:36	
Air Blank		0.000	01:37	
Subject Sample	# 1		01:39	Subject Test Refused
Breath Volume		0.000	01:39	
Air Blank		0.000	01:39	
Control Test		0.079	01:39	
Air Blank		0.000	01:40	
Diagnostics Check		OK	01:40	

Data Download Date/Time: 04/22/2016 18:25

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AMARO,RICARDO,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: SANCHEZ,GUILLERMO
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 03/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:15

SUBJECT'S LAST NAME: CORDOBA

DATE/TIME OF TEST: 04/14/2016 22:43:45

Diagnostics Check		OK	22:36	
Air Blank		0.000	22:36	
Control Test		0.079	22:37	
Air Blank		0.000	22:37	
Subject Sample	# 1	0.169	22:39	
Breath Volume		1.714	22:39	
Air Blank		0.000	22:39	
Air Blank		0.000	22:41	
Subject Sample	# 2	0.163	22:42	
Breath Volume		1.425	22:42	
Air Blank		0.000	22:42	
Control Test		0.074	22:43	Control Outside Tolerance
Air Blank		0.000	22:43	

Data Download Date/Time: 04/22/2016 18:25

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHERRELUS,RUDDENS,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: COVERT,SHAWN
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 03/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:42

SUBJECT'S LAST NAME: PRASS

DATE/TIME OF TEST: 04/15/2016 03:13:57

Diagnostics Check		OK	03:03
Air Blank		0.000	03:04
Control Test		0.079	03:04
Air Blank		0.000	03:04
Subject Sample	# 1	0.185	03:07
Breath Volume		1.265	03:07
Air Blank		0.000	03:08
Air Blank		0.000	03:10
Subject Sample	# 2	0.176	03:12
Breath Volume		1.355	03:12
Air Blank		0.000	03:13
Control Test		0.077	03:13
Air Blank		0.000	03:13
Diagnostics Check		OK	03:13

Data Download Date/Time: 04/22/2016 18:25

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GROSSMAN,JAY,J
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: GROSSMAN,JAY
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 03/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 05:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:50

SUBJECT'S LAST NAME: JIMENEZ

DATE/TIME OF TEST: 04/15/2016 06:14:14

Diagnostics Check		OK	06:11	
Air Blank		0.000	06:11	
Control Test		0.078	06:12	
Air Blank		0.000	06:12	
Subject Sample	# 1		06:12	Subject Test Refused
Breath Volume		0.000	06:12	
Air Blank		0.000	06:13	
Control Test		0.078	06:13	
Air Blank		0.000	06:14	
Diagnostics Check		OK	06:14	

Data Download Date/Time: 04/22/2016 18:25

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PASTOR,JOSE,F
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: PASTOR,JOSE
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 03/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 17:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:29

SUBJECT'S LAST NAME: VASQUEZ GOMEZ

DATE/TIME OF TEST: 04/15/2016 18:01:23

Diagnostics Check		OK	17:54
Air Blank		0.000	17:55
Control Test		0.078	17:55
Air Blank		0.000	17:56
Subject Sample	# 1	0.135	17:56
Breath Volume		3.523	17:56
Air Blank		0.000	17:57
Air Blank		0.000	17:59
Subject Sample	# 2	0.144	17:59
Breath Volume		3.519	17:59
Air Blank		0.000	18:00
Control Test		0.078	18:00
Air Blank		0.000	18:01
Diagnostics Check		OK	18:01

Data Download Date/Time: 04/22/2016 18:25

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MIRANDA,IVAN,L
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: MIRANDA,IVAN
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 03/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 04/15/2016 21:59:48

Diagnostics Check		OK	21:55	
Air Blank		0.000	21:55	
Control Test		0.079	21:56	
Air Blank		0.000	21:56	
Subject Sample	# 1		21:59	Sequence Aborted
Breath Volume		0.449	21:59	
Air Blank		0.000	21:59	

Data Download Date/Time: 04/22/2016 18:25

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MIRANDA,IVAN,L
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: MIRANDA,IVAN
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 03/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 04/15/2016 22:09:22

Diagnostics Check		OK	22:02	
Air Blank		0.000	22:02	
Control Test		0.076	22:02	
Air Blank		0.000	22:03	
Subject Sample	# 1	0.260	22:04	
Breath Volume		3.867	22:04	
Air Blank		0.000	22:05	
Air Blank		0.000	22:07	
Subject Sample	# 2	0.254	22:07	
Breath Volume		3.382	22:07	
Air Blank		0.000	22:08	
Control Test		0.072	22:08	Control Outside Tolerance
Air Blank		0.000	22:09	

Data Download Date/Time: 04/22/2016 18:25

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MIRANDA,IVAN,L
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: MIRANDA,IVAN
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 03/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 04/15/2016 22:28:27

Diagnostics Check		OK	22:21
Air Blank		0.000	22:22
Control Test		0.075	22:22
Air Blank		0.000	22:23
Subject Sample	# 1	0.253	22:23
Breath Volume		1.839	22:23
Air Blank		0.000	22:24
Air Blank		0.000	22:26
Subject Sample	# 2	0.250	22:26
Breath Volume		2.191	22:26
Air Blank		0.000	22:27
Control Test		0.077	22:27
Air Blank		0.000	22:28
Diagnostics Check		OK	22:28

Data Download Date/Time: 04/22/2016 18:25

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERNANDEZ,VANESSA,G
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: FERNANDEZ,VANESSA
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 03/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 07:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:30

SUBJECT'S LAST NAME: RODRIGUEZ BENCOMO

DATE/TIME OF TEST: 04/17/2016 09:07:33

Diagnostics Check		OK	08:55	
Air Blank		0.000	08:56	
Control Test		0.078	08:56	
Air Blank		0.000	08:57	
Subject Sample	# 1	0.158	09:00	Volume Not Met
Breath Volume		0.230	09:00	
Air Blank		0.000	09:01	
Air Blank		0.000	09:02	
Subject Sample	# 2	0.142	09:05	Volume Not Met
Breath Volume		0.136	09:05	
Air Blank		0.000	09:06	
Control Test		0.077	09:06	
Air Blank		0.000	09:07	
Diagnostics Check		OK	09:07	

Data Download Date/Time: 04/22/2016 18:25

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERNANDEZ,VANESSA,G
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: FERNANDEZ,VANESSA
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 03/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 07:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:30

SUBJECT'S LAST NAME: RODRIGUEZ BENCOMO

DATE/TIME OF TEST: 04/17/2016 09:18:13

Diagnostics Check		OK	09:14	
Air Blank		0.000	09:15	
Control Test		0.077	09:15	
Air Blank		0.000	09:16	
Subject Sample	# 1		09:16	Subject Test Refused
Breath Volume		0.175	09:16	
Air Blank		0.000	09:17	
Control Test		0.078	09:17	
Air Blank		0.000	09:18	
Diagnostics Check		OK	09:18	

Data Download Date/Time: 04/22/2016 18:25

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HEREDIA RUBIO,ROBERTO,A
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: HEREDIA RUBIO,ROBERTO
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 03/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 17:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:34

SUBJECT'S LAST NAME: SANTOS

DATE/TIME OF TEST: 04/17/2016 18:06:53

Diagnostics Check		OK	18:00
Air Blank		0.000	18:00
Control Test		0.078	18:01
Air Blank		0.000	18:01
Subject Sample	# 1	0.292	18:02
Breath Volume		2.125	18:02
Air Blank		0.000	18:03
Air Blank		0.000	18:04
Subject Sample	# 2	0.278	18:05
Breath Volume		1.214	18:05
Air Blank		0.000	18:05
Control Test		0.076	18:06
Air Blank		0.000	18:06
Diagnostics Check		OK	18:06

Data Download Date/Time: 04/22/2016 18:25

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,LUIS,G
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: VENTURA,AMADO
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 03/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:42

SUBJECT'S LAST NAME: CHAVEZ

DATE/TIME OF TEST: 04/18/2016 05:12:22

Diagnostics Check		OK	05:05
Air Blank		0.000	05:05
Control Test		0.078	05:05
Air Blank		0.000	05:06
Subject Sample	# 1	0.166	05:07
Breath Volume		1.128	05:07
Air Blank		0.000	05:08
Air Blank		0.000	05:09
Subject Sample	# 2	0.173	05:10
Breath Volume		1.605	05:10
Air Blank		0.000	05:11
Control Test		0.078	05:11
Air Blank		0.000	05:12
Diagnostics Check		OK	05:12

Data Download Date/Time: 04/22/2016 18:25

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAZARIEGOS,SANDRA,E
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: HERRERA,J
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 03/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:03

SUBJECT'S LAST NAME: MAZARIEGOS

DATE/TIME OF TEST: 04/19/2016 02:28:34

Diagnostics Check		OK	02:23	
Air Blank		0.000	02:24	
Control Test		0.078	02:24	
Air Blank		0.000	02:24	
Subject Sample	# 1		02:26	Subject Test Refused
Breath Volume		0.300	02:26	
Air Blank		0.000	02:27	
Control Test		0.079	02:27	
Air Blank		0.000	02:28	
Diagnostics Check		OK	02:28	

Data Download Date/Time: 04/22/2016 18:25

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,RYAN,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: RODRIGUEZ,RYAN
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 03/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 08:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:20

SUBJECT'S LAST NAME: AGUILERA

DATE/TIME OF TEST: 04/19/2016 08:54:30

Diagnostics Check		OK	08:48
Air Blank		0.000	08:48
Control Test		0.078	08:48
Air Blank		0.000	08:49
Subject Sample	# 1	0.189	08:50
Breath Volume		3.222	08:50
Air Blank		0.000	08:50
Air Blank		0.000	08:52
Subject Sample	# 2	0.183	08:52
Breath Volume		3.480	08:52
Air Blank		0.000	08:53
Control Test		0.078	08:53
Air Blank		0.000	08:54
Diagnostics Check		OK	08:54

Data Download Date/Time: 04/22/2016 18:25

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PASTOR,JOSE,F
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: PASTOR,JOSE
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 03/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:58

SUBJECT'S LAST NAME: MARTINEZ CHAGUENDO

DATE/TIME OF TEST: 04/20/2016 00:23:40

Diagnostics Check		OK	00:20	
Air Blank		0.000	00:21	
Control Test		0.079	00:21	
Air Blank		0.000	00:21	
Subject Sample	# 1		00:22	Subject Test Refused
Breath Volume		0.000	00:22	
Air Blank		0.000	00:22	
Control Test		0.079	00:23	
Air Blank		0.000	00:23	
Diagnostics Check		OK	00:23	

Data Download Date/Time: 04/22/2016 18:25

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PASTOR,JOSE,F
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: PASTOR,JOSE
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 03/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: SANTANA

DATE/TIME OF TEST: 04/21/2016 23:03:53

Diagnostics Check		OK	22:57
Air Blank		0.000	22:58
Control Test		0.078	22:58
Air Blank		0.000	22:58
Subject Sample	# 1	0.098	22:59
Breath Volume		2.890	22:59
Air Blank		0.000	23:00
Air Blank		0.000	23:01
Subject Sample	# 2	0.104	23:02
Breath Volume		3.500	23:02
Air Blank		0.000	23:02
Control Test		0.078	23:03
Air Blank		0.000	23:03
Diagnostics Check		OK	23:03

Data Download Date/Time: 04/22/2016 18:25

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHERRELUS,RUDDENS,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: CHERRELUS,RUDDENS
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 03/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:08

SUBJECT'S LAST NAME: MALDONADO

DATE/TIME OF TEST: 04/22/2016 05:35:00

Diagnostics Check		OK	05:28
Air Blank		0.000	05:29
Control Test		0.078	05:29
Air Blank		0.000	05:29
Subject Sample	# 1	0.185	05:30
Breath Volume		2.734	05:30
Air Blank		0.000	05:31
Air Blank		0.000	05:32
Subject Sample	# 2	0.180	05:33
Breath Volume		2.320	05:33
Air Blank		0.000	05:34
Control Test		0.079	05:34
Air Blank		0.000	05:34
Diagnostics Check		OK	05:35

Data Download Date/Time: 04/22/2016 18:25

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TADLOCK,JAYNIS,L
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: TADLOCK,JAYNIS
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 04/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:12

SUBJECT'S LAST NAME: WALTON

DATE/TIME OF TEST: 04/23/2016 06:55:07

Diagnostics Check		OK	06:48
Air Blank		0.000	06:48
Control Test		0.079	06:49
Air Blank		0.000	06:49
Subject Sample	# 1	0.059	06:50
Breath Volume		1.378	06:50
Air Blank		0.000	06:51
Air Blank		0.000	06:53
Subject Sample	# 2	0.060	06:53
Breath Volume		1.570	06:53
Air Blank		0.000	06:54
Control Test		0.077	06:54
Air Blank		0.000	06:55
Diagnostics Check		OK	06:55

Data Download Date/Time: 05/25/2016 19:12

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DECESPEDES,CARLOS,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: MILAN,RAFAEL
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 04/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 16:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:15

SUBJECT'S LAST NAME: BUCARDO

DATE/TIME OF TEST: 04/23/2016 17:41:28

Diagnostics Check		OK	17:35
Air Blank		0.000	17:35
Control Test		0.078	17:36
Air Blank		0.000	17:36
Subject Sample	# 1	0.299	17:37
Breath Volume		1.722	17:37
Air Blank		0.000	17:37
Air Blank		0.000	17:39
Subject Sample	# 2	0.306	17:39
Breath Volume		1.644	17:39
Air Blank		0.000	17:40
Control Test		0.078	17:40
Air Blank		0.000	17:41
Diagnostics Check		OK	17:41

Data Download Date/Time: 05/25/2016 19:12

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,JASIEL,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: SADLER,JENEVIVE
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 04/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:44

SUBJECT'S LAST NAME: FURLAN

DATE/TIME OF TEST: 04/24/2016 03:12:33

Diagnostics Check		OK	03:05
Air Blank		0.000	03:05
Control Test		0.079	03:05
Air Blank		0.000	03:06
Subject Sample	# 1	0.187	03:07
Breath Volume		2.871	03:07
Air Blank		0.000	03:07
Air Blank		0.000	03:09
Subject Sample	# 2	0.190	03:11
Breath Volume		1.988	03:11
Air Blank		0.000	03:11
Control Test		0.078	03:11
Air Blank		0.000	03:12
Diagnostics Check		OK	03:12

Data Download Date/Time: 05/25/2016 19:12

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERNANDEZ,VANESSA,G
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: FERNANDEZ,VANESSA
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 03/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 10:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:53

SUBJECT'S LAST NAME: SANCHEZ

DATE/TIME OF TEST: 04/09/2016 11:42:26

Diagnostics Check		OK	11:36
Air Blank		0.000	11:36
Control Test		0.079	11:36
Air Blank		0.000	11:37
Subject Sample	# 1	0.141	11:38
Breath Volume		1.109	11:38
Air Blank		0.000	11:38
Air Blank		0.000	11:40
Subject Sample	# 2	0.151	11:40
Breath Volume		1.101	11:40
Air Blank		0.000	11:41
Control Test		0.078	11:41
Air Blank		0.000	11:42
Diagnostics Check		OK	11:42

Data Download Date/Time: 04/22/2016 18:18

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MATHIS,ANDRE,D
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: CARMONA,R
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 03/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: SCHEIDEL

DATE/TIME OF TEST: 04/16/2016 00:07:10

Diagnostics Check		OK	00:03	
Air Blank		0.000	00:03	
Control Test		0.079	00:03	
Air Blank		0.000	00:04	
Subject Sample	# 1		00:05	Subject Test Refused
Breath Volume		0.000	00:05	
Air Blank		0.000	00:06	
Control Test		0.079	00:06	
Air Blank		0.000	00:07	
Diagnostics Check		OK	00:07	

Data Download Date/Time: 04/22/2016 18:18

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TADLOCK,JAYNIS,L
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: GONZALEZ,M
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 04/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:12

SUBJECT'S LAST NAME: WALTON

DATE/TIME OF TEST: 04/23/2016 05:49:56

Diagnostics Check		OK	05:32	
Air Blank		0.000	05:33	
Control Test		0.079	05:33	
Air Blank		0.000	05:33	
Subject Sample	# 1	0.079	05:37	Volume Not Met
Breath Volume		0.054	05:37	
Air Blank		0.000	05:37	
Air Blank		0.000	05:39	
Subject Sample	# 2	0.073	05:42	
Breath Volume		2.648	05:42	
Air Blank		0.000	05:43	
Air Blank		0.000	05:45	
Subject Sample	# 3	0.071	05:48	Volume Not Met
Breath Volume		0.015	05:48	
Air Blank		0.000	05:49	
Control Test		0.074	05:49	Control Outside Tolerance
Air Blank		0.000	05:49	

Data Download Date/Time: 05/25/2016 19:14

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TADLOCK,JAYNIS,L
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: GONZALEZ,M
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 04/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:12

SUBJECT'S LAST NAME: WALTON

DATE/TIME OF TEST: 04/23/2016 05:58:41

Diagnostics Check		OK	05:52	
Air Blank		0.000	05:52	
Control Test		0.075	05:52	
Air Blank		0.000	05:53	
Subject Sample	# 1	0.073	05:54	
Breath Volume		2.351	05:54	
Air Blank		0.000	05:55	
Air Blank		0.000	05:56	
Subject Sample	# 2	0.073	05:57	
Breath Volume		1.203	05:57	
Air Blank		0.000	05:57	
Control Test		0.074	05:58	Control Outside Tolerance
Air Blank		0.000	05:58	

Data Download Date/Time: 05/25/2016 19:14

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TADLOCK,JAYNIS,L
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: TADLOCK,JAYNIS
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 04/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:12

SUBJECT'S LAST NAME: WALTON

DATE/TIME OF TEST: 04/23/2016 06:04:46

Diagnostics Check	OK	06:03	
Air Blank	0.000	06:03	
Control Test	0.067	06:04	Control Outside Tolerance
Air Blank	0.000	06:04	

Data Download Date/Time: 05/25/2016 19:14

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TADLOCK,JAYNIS,L
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: TADLOCK,JAYNIS
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 04/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:12

SUBJECT'S LAST NAME: WALTON

DATE/TIME OF TEST: 04/23/2016 06:36:38

Diagnostics Check	OK	06:35	
Air Blank	0.000	06:35	
Control Test	0.074	06:36	Control Outside Tolerance
Air Blank	0.000	06:36	

Data Download Date/Time: 05/25/2016 19:15

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEDE,EDDY,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: FORD,ERVENS
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 04/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 06:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:56

SUBJECT'S LAST NAME: HOGGINS

DATE/TIME OF TEST: 04/24/2016 10:20:35

Diagnostics Check		OK	10:17	
Air Blank		0.000	10:18	
Control Test		0.079	10:18	
Air Blank		0.000	10:19	
Subject Sample	# 1		10:19	Subject Test Refused
Breath Volume		0.000	10:19	
Air Blank		0.000	10:19	
Control Test		0.079	10:20	
Air Blank		0.000	10:20	
Diagnostics Check		OK	10:20	

Data Download Date/Time: 05/25/2016 19:15

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINEZ,MATTHEW,D
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: PALOMINO,JULIO
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 04/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: BAUTISTA

DATE/TIME OF TEST: 04/29/2016 03:33:08

Diagnostics Check		OK	03:25
Air Blank		0.000	03:26
Control Test		0.079	03:26
Air Blank		0.000	03:26
Subject Sample	# 1	0.216	03:28
Breath Volume		1.226	03:28
Air Blank		0.000	03:28
Air Blank		0.000	03:30
Subject Sample	# 2	0.220	03:31
Breath Volume		1.464	03:31
Air Blank		0.000	03:32
Control Test		0.077	03:32
Air Blank		0.000	03:33
Diagnostics Check		OK	03:33

Data Download Date/Time: 05/25/2016 19:15

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AMARO,RICARDO,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: RODRIGUEZ,LEILANY
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 04/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 15:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:18

SUBJECT'S LAST NAME: ALONSO

DATE/TIME OF TEST: 04/29/2016 15:44:59

Diagnostics Check		OK	15:38
Air Blank		0.000	15:39
Control Test		0.080	15:39
Air Blank		0.000	15:40
Subject Sample	# 1	0.139	15:40
Breath Volume		2.730	15:40
Air Blank		0.000	15:41
Air Blank		0.000	15:43
Subject Sample	# 2	0.134	15:43
Breath Volume		2.542	15:43
Air Blank		0.000	15:44
Control Test		0.077	15:44
Air Blank		0.000	15:44
Diagnostics Check		OK	15:44

Data Download Date/Time: 05/25/2016 19:15

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALOMINO,JULIO,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: PALOMINO,JULIO
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 04/22/2016

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: PEREZ DE DIOS

DATE/TIME OF TEST: 04/30/2016 00:23:01

Diagnostics Check		OK	00:14
Air Blank		0.000	00:15
Control Test		0.079	00:15
Air Blank		0.000	00:15
Subject Sample	# 1	0.190	00:17
Breath Volume		1.453	00:17
Air Blank		0.000	00:18
Air Blank		0.000	00:20
Subject Sample	# 2	0.179	00:21
Breath Volume		1.378	00:21
Air Blank		0.000	00:22
Control Test		0.078	00:22
Air Blank		0.000	00:22
Diagnostics Check		OK	00:23

Data Download Date/Time: 05/25/2016 19:15

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALOMINO,JULIO,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: PALOMINO,JULIO
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 04/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: PEREZ DE DIOS

DATE/TIME OF TEST: 04/30/2016 01:04:09

Diagnostics Check		OK	00:49	
Air Blank		0.000	00:49	
Control Test		0.079	00:50	
Air Blank		0.000	00:50	
Subject Sample	# 1	0.185	00:51	
Breath Volume		1.378	00:51	
Air Blank		0.000	00:52	
Air Blank		0.000	00:53	
Subject Sample	# 2		00:57	No Sample Provided
Breath Volume		0.000	00:57	
Air Blank		0.000	00:57	
Air Blank		0.000	00:59	
Subject Sample	# 3		01:02	No Sample Provided
Breath Volume		0.000	01:02	
Air Blank		0.000	01:03	
Control Test		0.079	01:03	
Air Blank		0.000	01:04	
Diagnostics Check		OK	01:04	

Data Download Date/Time: 05/25/2016 19:15

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CABEZA,HIRAM,E
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: CABEZA,HIRAM
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 04/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 16:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:55

SUBJECT'S LAST NAME: BARRERA CABALLERO

DATE/TIME OF TEST: 04/30/2016 17:24:46

Diagnostics Check		OK	17:21	
Air Blank		0.000	17:21	
Control Test		0.078	17:21	
Air Blank		0.000	17:22	
Subject Sample	# 1		17:23	Subject Test Refused
Breath Volume		0.000	17:23	
Air Blank		0.000	17:23	
Control Test		0.078	17:24	
Air Blank		0.000	17:24	
Diagnostics Check		OK	17:24	

Data Download Date/Time: 05/25/2016 19:15

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARNANDAN,INDROWTIE,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: MANNING,MARK
ARRESTING OFFICER AGENCY: Pembroke Pines PD
LAST AGENCY INSPECTION: 03/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 05:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:40

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 04/01/2016 06:08:01

Diagnostics Check		OK	06:01
Air Blank		0.000	06:01
Control Test		0.080	06:02
Air Blank		0.000	06:02
Subject Sample	# 1	0.152	06:03
Breath Volume		1.117	06:03
Air Blank		0.000	06:04
Air Blank		0.000	06:05
Subject Sample	# 2	0.158	06:06
Breath Volume		1.976	06:06
Air Blank		0.000	06:07
Control Test		0.079	06:07
Air Blank		0.000	06:07
Diagnostics Check		OK	06:08

Data Download Date/Time: 04/25/2016 15:08

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SAPP,JOSHUA
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 03/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 23:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:19

SUBJECT'S LAST NAME: IRVINE

DATE/TIME OF TEST: 04/02/2016 01:05:08

Diagnostics Check		OK	00:53	
Air Blank		0.000	00:53	
Control Test		0.079	00:54	
Air Blank		0.000	00:54	
Subject Sample	# 1	0.160	00:57	Volume Not Met
Breath Volume		0.093	00:57	
Air Blank		0.000	00:58	
Air Blank		0.000	01:00	
Subject Sample	# 2	0.141	01:03	Volume Not Met
Breath Volume		0.164	01:03	
Air Blank		0.000	01:04	
Control Test		0.078	01:04	
Air Blank		0.000	01:05	
Diagnostics Check		OK	01:05	

Data Download Date/Time: 04/25/2016 15:08

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SAPP,JOSHUA
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 03/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 23:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:19

SUBJECT'S LAST NAME: IRVINE

DATE/TIME OF TEST: 04/02/2016 01:21:56

Diagnostics Check		OK	01:07	
Air Blank		0.000	01:07	
Control Test		0.077	01:07	
Air Blank		0.000	01:08	
Subject Sample	# 1	0.189	01:09	
Breath Volume		2.523	01:09	
Air Blank		0.000	01:09	
Air Blank		0.000	01:11	
Subject Sample	# 2	0.112	01:14	Volume Not Met
Breath Volume		0.757	01:14	
Air Blank		0.000	01:15	
Air Blank		0.000	01:17	
Subject Sample	# 3	0.108	01:20	Volume Not Met
Breath Volume		0.265	01:20	
Air Blank		0.000	01:20	
Control Test		0.078	01:21	
Air Blank		0.000	01:21	
Diagnostics Check		OK	01:21	

Data Download Date/Time: 04/25/2016 15:08

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: PERCIVAL,KAITLYN
ARRESTING OFFICER AGENCY: Pembroke Pines PD
LAST AGENCY INSPECTION: 03/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 02:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: LOZANO

DATE/TIME OF TEST: 04/02/2016 03:22:29

Diagnostics Check		OK	03:09	
Air Blank		0.000	03:09	
Control Test		0.079	03:10	
Air Blank		0.000	03:10	
Subject Sample	# 1	0.170	03:13	Volume Not Met
Breath Volume		0.019	03:13	
Air Blank		0.000	03:14	
Air Blank		0.000	03:16	
Subject Sample	# 2	0.181	03:17	
Breath Volume		1.386	03:17	
Air Blank		0.000	03:18	
Air Blank		0.000	03:20	
Subject Sample	# 3	0.183	03:20	
Breath Volume		2.082	03:20	
Air Blank		0.000	03:21	
Control Test		0.076	03:21	
Air Blank		0.000	03:22	
Diagnostics Check		OK	03:22	

Data Download Date/Time: 04/25/2016 15:08

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: MARRERO,DAVID
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 03/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 01:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:48

SUBJECT'S LAST NAME: ORTIZ SEVILLA

DATE/TIME OF TEST: 04/03/2016 02:27:05

Diagnostics Check		OK	02:15	
Air Blank		0.000	02:15	
Control Test		0.079	02:16	
Air Blank		0.000	02:16	
Subject Sample	# 1	0.182	02:19	Volume Not Met
Breath Volume		0.187	02:19	
Air Blank		0.000	02:20	
Air Blank		0.000	02:22	
Subject Sample	# 2	0.149	02:25	Volume Not Met
Breath Volume		0.207	02:25	
Air Blank		0.000	02:26	
Control Test		0.078	02:26	
Air Blank		0.000	02:27	
Diagnostics Check		OK	02:27	

Data Download Date/Time: 04/25/2016 15:08

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: PRESUME,JEHUD
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 03/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 02:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:26

SUBJECT'S LAST NAME: CIFUENTES HERNANDEZ

DATE/TIME OF TEST: 04/03/2016 04:11:56

Diagnostics Check		OK	04:02
Air Blank		0.000	04:03
Control Test		0.079	04:03
Air Blank		0.000	04:04
Subject Sample	# 1	0.146	04:05
Breath Volume		1.191	04:05
Air Blank		0.000	04:06
Air Blank		0.000	04:08
Subject Sample	# 2	0.145	04:10
Breath Volume		2.738	04:10
Air Blank		0.000	04:11
Control Test		0.078	04:11
Air Blank		0.000	04:11
Diagnostics Check		OK	04:11

Data Download Date/Time: 04/25/2016 15:08

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: REYES,ROLANDO
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 03/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 17:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:30

SUBJECT'S LAST NAME: GARZON

DATE/TIME OF TEST: 04/03/2016 19:12:15

Diagnostics Check		OK	19:04
Air Blank		0.000	19:04
Control Test		0.080	19:04
Air Blank		0.000	19:05
Subject Sample	# 1	0.155	19:06
Breath Volume		1.609	19:06
Air Blank		0.000	19:06
Air Blank		0.000	19:08
Subject Sample	# 2	0.169	19:10
Breath Volume		1.820	19:10
Air Blank		0.000	19:11
Control Test		0.081	19:11
Air Blank		0.000	19:12
Diagnostics Check		OK	19:12

Data Download Date/Time: 04/25/2016 15:08

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: LLOYD,WAYNE
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 03/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 02:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: FELISMA

DATE/TIME OF TEST: 04/04/2016 03:18:17

Diagnostics Check		OK	03:11
Air Blank		0.000	03:11
Control Test		0.079	03:11
Air Blank		0.000	03:12
Subject Sample	# 1	0.185	03:13
Breath Volume		1.621	03:13
Air Blank		0.000	03:14
Air Blank		0.000	03:15
Subject Sample	# 2	0.167	03:16
Breath Volume		1.269	03:16
Air Blank		0.000	03:17
Control Test		0.078	03:17
Air Blank		0.000	03:18
Diagnostics Check		OK	03:18

Data Download Date/Time: 04/25/2016 15:08

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: WASHINGTON,YVETTE
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 03/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 19:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:40

SUBJECT'S LAST NAME: SAYKALY

DATE/TIME OF TEST: 04/07/2016 21:10:02

Diagnostics Check		OK	21:03
Air Blank		0.000	21:03
Control Test		0.080	21:03
Air Blank		0.000	21:04
Subject Sample	# 1	0.262	21:05
Breath Volume		2.519	21:05
Air Blank		0.000	21:06
Air Blank		0.000	21:07
Subject Sample	# 2	0.253	21:08
Breath Volume		2.121	21:08
Air Blank		0.000	21:09
Control Test		0.080	21:09
Air Blank		0.000	21:09
Diagnostics Check		OK	21:10

Data Download Date/Time: 04/25/2016 15:08

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARNANDAN,INDROWTIE,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: PRESUME,JEHUD
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 03/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 04:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:25

SUBJECT'S LAST NAME: SHINE

DATE/TIME OF TEST: 04/09/2016 05:51:43

Diagnostics Check		OK	05:45
Air Blank		0.000	05:45
Control Test		0.080	05:46
Air Blank		0.000	05:46
Subject Sample	# 1	0.115	05:47
Breath Volume		1.289	05:47
Air Blank		0.000	05:47
Air Blank		0.000	05:49
Subject Sample	# 2	0.112	05:50
Breath Volume		2.187	05:50
Air Blank		0.000	05:50
Control Test		0.079	05:51
Air Blank		0.000	05:51
Diagnostics Check		OK	05:51

Data Download Date/Time: 04/25/2016 15:08

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINA,MARMILLY,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: PRESUME,JEHUD
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 03/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: CARDONA

DATE/TIME OF TEST: 04/10/2016 02:40:45

Diagnostics Check		OK	02:33
Air Blank		0.000	02:34
Control Test		0.080	02:34
Air Blank		0.000	02:34
Subject Sample	# 1	0.245	02:35
Breath Volume		1.730	02:35
Air Blank		0.000	02:36
Air Blank		0.000	02:38
Subject Sample	# 2	0.253	02:39
Breath Volume		1.214	02:39
Air Blank		0.000	02:39
Control Test		0.079	02:40
Air Blank		0.000	02:40
Diagnostics Check		OK	02:40

Data Download Date/Time: 04/25/2016 15:08

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARNANDAN,INDROWTIE,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: DANG,CHRIS
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 03/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 08:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:02

SUBJECT'S LAST NAME: MALDONADO SANCHEZ

DATE/TIME OF TEST: 04/10/2016 09:32:51

Diagnostics Check		OK	09:26
Air Blank		0.000	09:27
Control Test		0.081	09:27
Air Blank		0.000	09:27
Subject Sample	# 1	0.150	09:28
Breath Volume		1.269	09:28
Air Blank		0.000	09:29
Air Blank		0.000	09:30
Subject Sample	# 2	0.148	09:31
Breath Volume		1.164	09:31
Air Blank		0.000	09:31
Control Test		0.080	09:32
Air Blank		0.000	09:32
Diagnostics Check		OK	09:32

Data Download Date/Time: 04/25/2016 15:08

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: DEDEJ,MEDJON
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 03/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 20:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:34

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 04/10/2016 22:05:24

Diagnostics Check		OK	21:58
Air Blank		0.000	21:59
Control Test		0.080	21:59
Air Blank		0.000	22:00
Subject Sample	# 1	0.218	22:00
Breath Volume		2.191	22:00
Air Blank		0.000	22:01
Air Blank		0.000	22:03
Subject Sample	# 2	0.201	22:03
Breath Volume		1.515	22:03
Air Blank		0.000	22:04
Control Test		0.080	22:04
Air Blank		0.000	22:05
Diagnostics Check		OK	22:05

Data Download Date/Time: 04/25/2016 15:08

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: COFFIN,BRIAN
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 03/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 18:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:55

SUBJECT'S LAST NAME: PELLERITO

DATE/TIME OF TEST: 04/12/2016 20:25:19

Diagnostics Check		OK	20:18
Air Blank		0.000	20:18
Control Test		0.080	20:19
Air Blank		0.000	20:19
Subject Sample	# 1	0.067	20:20
Breath Volume		2.156	20:20
Air Blank		0.000	20:20
Air Blank		0.000	20:22
Subject Sample	# 2	0.066	20:23
Breath Volume		3.695	20:23
Air Blank		0.000	20:24
Control Test		0.080	20:24
Air Blank		0.000	20:25
Diagnostics Check		OK	20:25

Data Download Date/Time: 04/25/2016 15:08

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARNANDAN,INDROWTIE,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SAINT-AMOUR,JEAN
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 03/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 00:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: MOLINA

DATE/TIME OF TEST: 04/14/2016 01:22:18

Diagnostics Check		OK	01:15
Air Blank		0.000	01:16
Control Test		0.079	01:16
Air Blank		0.000	01:17
Subject Sample	# 1	0.185	01:17
Breath Volume		1.710	01:17
Air Blank		0.000	01:18
Air Blank		0.000	01:20
Subject Sample	# 2	0.183	01:20
Breath Volume		2.031	01:20
Air Blank		0.000	01:21
Control Test		0.078	01:21
Air Blank		0.000	01:22
Diagnostics Check		OK	01:22

Data Download Date/Time: 04/25/2016 15:08

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARNANDAN,INDROWTIE,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: PLUTA,DARIUSZ
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 03/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 01:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:52

SUBJECT'S LAST NAME: JERKINS

DATE/TIME OF TEST: 04/15/2016 02:20:21

Diagnostics Check		OK	02:14
Air Blank		0.000	02:14
Control Test		0.080	02:14
Air Blank		0.000	02:15
Subject Sample	# 1	0.150	02:15
Breath Volume		1.488	02:15
Air Blank		0.000	02:16
Air Blank		0.000	02:18
Subject Sample	# 2	0.138	02:18
Breath Volume		2.519	02:18
Air Blank		0.000	02:19
Control Test		0.080	02:19
Air Blank		0.000	02:20
Diagnostics Check		OK	02:20

Data Download Date/Time: 04/25/2016 15:08

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SERPHOS,STEVEN
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 03/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 00:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:06

SUBJECT'S LAST NAME: WINKLEMAN

DATE/TIME OF TEST: 04/16/2016 01:41:37

Diagnostics Check		OK	01:35
Air Blank		0.000	01:35
Control Test		0.080	01:36
Air Blank		0.000	01:36
Subject Sample	# 1	0.153	01:37
Breath Volume		1.945	01:37
Air Blank		0.000	01:37
Air Blank		0.000	01:39
Subject Sample	# 2	0.154	01:39
Breath Volume		2.460	01:39
Air Blank		0.000	01:40
Control Test		0.079	01:40
Air Blank		0.000	01:41
Diagnostics Check		OK	01:41

Data Download Date/Time: 04/25/2016 15:08

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARNANDAN,INDROWTIE,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ALBIZU,LEO
ARRESTING OFFICER AGENCY: Margate PD
LAST AGENCY INSPECTION: 03/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 06:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:18

SUBJECT'S LAST NAME: NATAL

DATE/TIME OF TEST: 04/16/2016 07:45:05

Diagnostics Check		OK	07:38
Air Blank		0.000	07:38
Control Test		0.079	07:38
Air Blank		0.000	07:39
Subject Sample	# 1	0.153	07:40
Breath Volume		1.945	07:40
Air Blank		0.000	07:40
Air Blank		0.000	07:42
Subject Sample	# 2	0.150	07:43
Breath Volume		1.687	07:43
Air Blank		0.000	07:44
Control Test		0.080	07:44
Air Blank		0.000	07:45
Diagnostics Check		OK	07:45

Data Download Date/Time: 04/25/2016 15:08

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: WASHINGTON,YVETTE
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 03/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 02:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: SADAUSKAS

DATE/TIME OF TEST: 04/17/2016 04:03:22

Diagnostics Check		OK	03:56
Air Blank		0.000	03:56
Control Test		0.078	03:56
Air Blank		0.000	03:57
Subject Sample	# 1	0.163	03:58
Breath Volume		3.035	03:58
Air Blank		0.000	03:58
Air Blank		0.000	04:00
Subject Sample	# 2	0.164	04:01
Breath Volume		3.050	04:01
Air Blank		0.000	04:02
Control Test		0.078	04:02
Air Blank		0.000	04:03
Diagnostics Check		OK	04:03

Data Download Date/Time: 04/25/2016 15:08

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ECKINGER,LANI,A
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: REYES,ROLANDO
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 03/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 11:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 12:54

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 04/18/2016 13:24:52

Diagnostics Check		OK	13:18
Air Blank		0.000	13:18
Control Test		0.080	13:18
Air Blank		0.000	13:19
Subject Sample	# 1	0.000	13:20
Breath Volume		1.199	13:20
Air Blank		0.000	13:20
Air Blank		0.000	13:22
Subject Sample	# 2	0.000	13:23
Breath Volume		1.261	13:23
Air Blank		0.000	13:23
Control Test		0.080	13:24
Air Blank		0.000	13:24
Diagnostics Check		OK	13:24

Data Download Date/Time: 04/25/2016 15:08

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: HORNE,PHILLIP
ARRESTING OFFICER AGENCY: Margate PD
LAST AGENCY INSPECTION: 03/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 18:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:45

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 04/18/2016 20:20:17

Diagnostics Check		OK	20:11	
Air Blank		0.000	20:12	
Control Test		0.079	20:12	
Air Blank		0.000	20:12	
Subject Sample	# 1	0.000	20:16	Volume Not Met
Breath Volume		0.128	20:16	
Air Blank		0.000	20:16	
Air Blank		0.000	20:18	
Subject Sample	# 2		20:18	Subject Test Refused
Breath Volume		0.000	20:18	
Air Blank		0.000	20:19	
Control Test		0.078	20:19	
Air Blank		0.000	20:20	
Diagnostics Check		OK	20:20	

Data Download Date/Time: 04/25/2016 15:08

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINA,MARMILLY,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: POPICK,SCOTT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 03/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 22:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:57

SUBJECT'S LAST NAME: CRAIG

DATE/TIME OF TEST: 04/18/2016 23:38:18

Diagnostics Check		OK	23:31
Air Blank		0.000	23:31
Control Test		0.080	23:31
Air Blank		0.000	23:32
Subject Sample	# 1	0.101	23:33
Breath Volume		1.878	23:33
Air Blank		0.000	23:34
Air Blank		0.000	23:36
Subject Sample	# 2	0.101	23:36
Breath Volume		1.601	23:36
Air Blank		0.000	23:37
Control Test		0.081	23:37
Air Blank		0.000	23:38
Diagnostics Check		OK	23:38

Data Download Date/Time: 04/25/2016 15:08

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARNANDAN,INDROWTIE,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: DALTON,RYAN
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 03/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 23:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:08

SUBJECT'S LAST NAME: GOETZ

DATE/TIME OF TEST: 04/20/2016 00:35:17

Diagnostics Check		OK	00:28
Air Blank		0.000	00:29
Control Test		0.079	00:29
Air Blank		0.000	00:30
Subject Sample	# 1	0.334	00:30
Breath Volume		1.664	00:30
Air Blank		0.000	00:31
Air Blank		0.000	00:33
Subject Sample	# 2	0.330	00:33
Breath Volume		1.535	00:33
Air Blank		0.000	00:34
Control Test		0.078	00:34
Air Blank		0.000	00:35
Diagnostics Check		OK	00:35

Data Download Date/Time: 04/25/2016 15:08

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: POPICK,SCOTT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 03/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 00:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: RAMIREZ

DATE/TIME OF TEST: 04/21/2016 01:40:08

Diagnostics Check		OK	01:36	
Air Blank		0.000	01:37	
Control Test		0.079	01:37	
Air Blank		0.000	01:38	
Subject Sample	# 1		01:38	Slope Not Met
Breath Volume		0.882	01:38	
Air Blank		0.000	01:39	
Control Test		0.079	01:39	
Air Blank		0.000	01:40	
Diagnostics Check		OK	01:40	

Data Download Date/Time: 04/25/2016 15:08

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: POPICK,SCOTT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 03/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 00:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:42

SUBJECT'S LAST NAME: RAMIREZ

DATE/TIME OF TEST: 04/21/2016 02:08:42

Diagnostics Check		OK	02:02
Air Blank		0.000	02:02
Control Test		0.079	02:03
Air Blank		0.000	02:03
Subject Sample	# 1	0.238	02:04
Breath Volume		1.246	02:04
Air Blank		0.000	02:04
Air Blank		0.000	02:06
Subject Sample	# 2	0.234	02:07
Breath Volume		1.339	02:07
Air Blank		0.000	02:07
Control Test		0.078	02:08
Air Blank		0.000	02:08
Diagnostics Check		OK	02:08

Data Download Date/Time: 04/25/2016 15:08

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARNANDAN,INDROWTIE,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: BURKE,BENJAMIN
ARRESTING OFFICER AGENCY: Pembroke Pines PD
LAST AGENCY INSPECTION: 03/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 03:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:18

SUBJECT'S LAST NAME: BENJAMIN

DATE/TIME OF TEST: 04/21/2016 04:45:43

Diagnostics Check		OK	04:38
Air Blank		0.000	04:38
Control Test		0.079	04:39
Air Blank		0.000	04:39
Subject Sample	# 1	0.195	04:41
Breath Volume		3.207	04:41
Air Blank		0.000	04:41
Air Blank		0.000	04:43
Subject Sample	# 2	0.200	04:44
Breath Volume		1.765	04:44
Air Blank		0.000	04:44
Control Test		0.077	04:45
Air Blank		0.000	04:45
Diagnostics Check		OK	04:45

Data Download Date/Time: 04/25/2016 15:08

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARNANDAN,INDROWTIE,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: BURKE,BENJAMIN
ARRESTING OFFICER AGENCY: Pembroke Pines PD
LAST AGENCY INSPECTION: 03/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 02:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:58

SUBJECT'S LAST NAME: MATA

DATE/TIME OF TEST: 04/22/2016 03:26:42

Diagnostics Check		OK	03:20
Air Blank		0.000	03:20
Control Test		0.079	03:21
Air Blank		0.000	03:21
Subject Sample	# 1	0.189	03:22
Breath Volume		2.789	03:22
Air Blank		0.000	03:22
Air Blank		0.000	03:24
Subject Sample	# 2	0.183	03:25
Breath Volume		2.742	03:25
Air Blank		0.000	03:25
Control Test		0.079	03:26
Air Blank		0.000	03:26
Diagnostics Check		OK	03:26

Data Download Date/Time: 04/25/2016 15:08

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARNANDAN,INDROWTIE,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: HARRIS,MILTON
ARRESTING OFFICER AGENCY: Pembroke Pines PD
LAST AGENCY INSPECTION: 03/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 05:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:45

SUBJECT'S LAST NAME: CASTRO ROMERO

DATE/TIME OF TEST: 04/22/2016 06:13:39

Diagnostics Check		OK	06:06
Air Blank		0.000	06:07
Control Test		0.080	06:07
Air Blank		0.000	06:08
Subject Sample	# 1	0.184	06:08
Breath Volume		2.019	06:08
Air Blank		0.000	06:09
Air Blank		0.000	06:11
Subject Sample	# 2	0.179	06:12
Breath Volume		1.808	06:12
Air Blank		0.000	06:12
Control Test		0.079	06:13
Air Blank		0.000	06:13
Diagnostics Check		OK	06:13

Data Download Date/Time: 04/25/2016 15:08

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: WEISSGERBER,MARK
ARRESTING OFFICER AGENCY: Davie PD
LAST AGENCY INSPECTION: 03/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 21:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: MEDINA CRUZ

DATE/TIME OF TEST: 04/23/2016 23:04:28

Diagnostics Check		OK	22:57
Air Blank		0.000	22:57
Control Test		0.080	22:57
Air Blank		0.000	22:58
Subject Sample	# 1	0.203	22:59
Breath Volume		1.507	22:59
Air Blank		0.000	23:00
Air Blank		0.000	23:02
Subject Sample	# 2	0.218	23:02
Breath Volume		1.496	23:02
Air Blank		0.000	23:03
Control Test		0.080	23:03
Air Blank		0.000	23:04
Diagnostics Check		OK	23:04

Data Download Date/Time: 04/25/2016 15:08

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: STRZALKOWSKI,SEAN
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 03/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 02:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:53

SUBJECT'S LAST NAME: MCINTOSH

DATE/TIME OF TEST: 04/24/2016 03:29:57

Diagnostics Check		OK	03:13	
Air Blank		0.000	03:13	
Control Test		0.079	03:14	
Air Blank		0.000	03:14	
Subject Sample	# 1	0.169	03:17	Volume Not Met
Breath Volume		0.449	03:17	
Air Blank		0.000	03:18	
Air Blank		0.000	03:20	
Subject Sample	# 2	0.168	03:22	
Breath Volume		1.316	03:22	
Air Blank		0.000	03:23	
Air Blank		0.000	03:25	
Subject Sample	# 3	0.174	03:28	Volume Not Met
Breath Volume		0.152	03:28	
Air Blank		0.000	03:28	
Control Test		0.079	03:29	
Air Blank		0.000	03:29	
Diagnostics Check		OK	03:29	

Data Download Date/Time: 04/25/2016 15:08

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINA,MARMILLY,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: LUMPKINS,HENRY
ARRESTING OFFICER AGENCY: Hollywood PD
LAST AGENCY INSPECTION: 03/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 02:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: KHAN

DATE/TIME OF TEST: 04/24/2016 03:41:44

Diagnostics Check		OK	03:35
Air Blank		0.000	03:35
Control Test		0.078	03:36
Air Blank		0.000	03:36
Subject Sample	# 1	0.106	03:37
Breath Volume		2.777	03:37
Air Blank		0.000	03:37
Air Blank		0.000	03:39
Subject Sample	# 2	0.106	03:40
Breath Volume		2.863	03:40
Air Blank		0.000	03:40
Control Test		0.079	03:41
Air Blank		0.000	03:41
Diagnostics Check		OK	03:41

Data Download Date/Time: 04/25/2016 15:09

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINA,MARMILLY,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: PERCIVAL,KAITLYN
ARRESTING OFFICER AGENCY: Pembroke Pines PD
LAST AGENCY INSPECTION: 03/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 06:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:15

SUBJECT'S LAST NAME: CETO CABO

DATE/TIME OF TEST: 04/24/2016 07:43:56

Diagnostics Check		OK	07:37
Air Blank		0.000	07:38
Control Test		0.080	07:38
Air Blank		0.000	07:38
Subject Sample	# 1	0.051	07:39
Breath Volume		2.421	07:39
Air Blank		0.000	07:40
Air Blank		0.000	07:41
Subject Sample	# 2	0.050	07:42
Breath Volume		2.367	07:42
Air Blank		0.000	07:43
Control Test		0.079	07:43
Air Blank		0.000	07:43
Diagnostics Check		OK	07:43

Data Download Date/Time: 04/25/2016 15:09

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DORSEY,PAMELA,F
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: PUJOL,JESUS
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 03/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A3
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:22

SUBJECT'S LAST NAME: SAYKALY

DATE/TIME OF TEST: 04/01/2016 00:56:03

Diagnostics Check		OK	00:48
Air Blank		0.000	00:48
Control Test		0.076	00:49
Air Blank		0.000	00:49
Subject Sample	# 1	0.257	00:50
Breath Volume		2.609	00:50
Air Blank		0.000	00:51
Air Blank		0.000	00:53
Subject Sample	# 2	0.242	00:54
Breath Volume		1.726	00:54
Air Blank		0.000	00:55
Control Test		0.077	00:55
Air Blank		0.000	00:55
Diagnostics Check		OK	00:56

Data Download Date/Time: 04/15/2016 13:04

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARI,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITE,EDWARD
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 03/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A3
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: DOMINGUEZ

DATE/TIME OF TEST: 04/01/2016 03:49:43

Diagnostics Check		OK	03:43
Air Blank		0.000	03:43
Control Test		0.078	03:43
Air Blank		0.000	03:44
Subject Sample	# 1	0.144	03:45
Breath Volume		3.328	03:45
Air Blank		0.000	03:45
Air Blank		0.000	03:47
Subject Sample	# 2	0.139	03:47
Breath Volume		2.898	03:47
Air Blank		0.000	03:48
Control Test		0.079	03:49
Air Blank		0.000	03:49
Diagnostics Check		OK	03:49

Data Download Date/Time: 04/15/2016 13:04

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DORSEY,PAMELA,F
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: DIEZ,DARRELL
ARRESTING OFFICER AGENCY: Palm Springs PD
LAST AGENCY INSPECTION: 03/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A3
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 12:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:41

SUBJECT'S LAST NAME: DE ORNELAS ROJAS

DATE/TIME OF TEST: 04/02/2016 01:18:05

Diagnostics Check		OK	01:10
Air Blank		0.000	01:10
Control Test		0.079	01:11
Air Blank		0.000	01:11
Subject Sample	# 1	0.160	01:13
Breath Volume		1.945	01:13
Air Blank		0.000	01:14
Air Blank		0.000	01:15
Subject Sample	# 2	0.154	01:16
Breath Volume		1.570	01:16
Air Blank		0.000	01:17
Control Test		0.077	01:17
Air Blank		0.000	01:18
Diagnostics Check		OK	01:18

Data Download Date/Time: 04/15/2016 13:04

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CHIQUITO-RODRIGUEZ,LENNDY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 03/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A3
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: MCMANUS

DATE/TIME OF TEST: 04/02/2016 23:20:12

Diagnostics Check		OK	23:12
Air Blank		0.000	23:12
Control Test		0.080	23:12
Air Blank		0.000	23:13
Subject Sample	# 1	0.056	23:15
Breath Volume		1.281	23:15
Air Blank		0.000	23:16
Air Blank		0.000	23:17
Subject Sample	# 2	0.056	23:18
Breath Volume		1.601	23:18
Air Blank		0.000	23:19
Control Test		0.078	23:19
Air Blank		0.000	23:20
Diagnostics Check		OK	23:20

Data Download Date/Time: 04/15/2016 13:04

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: AVEY,ADAM
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 03/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A3
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:17

SUBJECT'S LAST NAME: GOODWILL

DATE/TIME OF TEST: 04/03/2016 04:52:49

Diagnostics Check		OK	04:45
Air Blank		0.000	04:46
Control Test		0.080	04:46
Air Blank		0.000	04:47
Subject Sample	# 1	0.085	04:47
Breath Volume		1.777	04:47
Air Blank		0.000	04:48
Air Blank		0.000	04:50
Subject Sample	# 2	0.087	04:50
Breath Volume		1.824	04:50
Air Blank		0.000	04:51
Control Test		0.079	04:52
Air Blank		0.000	04:52
Diagnostics Check		OK	04:52

Data Download Date/Time: 04/15/2016 13:04

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FERNANDEZ,RYAN
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 03/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A3
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:45

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 04/03/2016 05:37:12

Diagnostics Check		OK	05:30
Air Blank		0.000	05:30
Control Test		0.080	05:30
Air Blank		0.000	05:31
Subject Sample	# 1	0.140	05:32
Breath Volume		2.171	05:32
Air Blank		0.000	05:33
Air Blank		0.000	05:34
Subject Sample	# 2	0.142	05:35
Breath Volume		2.382	05:35
Air Blank		0.000	05:36
Control Test		0.079	05:36
Air Blank		0.000	05:37
Diagnostics Check		OK	05:37

Data Download Date/Time: 04/15/2016 13:04

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WARD,CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 03/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A3
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:52

SUBJECT'S LAST NAME: SIEGEL

DATE/TIME OF TEST: 04/06/2016 04:29:06

Diagnostics Check		OK	04:21
Air Blank		0.000	04:21
Control Test		0.080	04:21
Air Blank		0.000	04:22
Subject Sample	# 1	0.211	04:23
Breath Volume		1.554	04:23
Air Blank		0.000	04:24
Air Blank		0.000	04:26
Subject Sample	# 2	0.207	04:27
Breath Volume		1.640	04:27
Air Blank		0.000	04:28
Control Test		0.078	04:28
Air Blank		0.000	04:29
Diagnostics Check		OK	04:29

Data Download Date/Time: 04/15/2016 13:04

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAIN,JEANETTE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: HOPE,PAUL
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 03/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A3
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 17:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:11

SUBJECT'S LAST NAME: BRAGDON

DATE/TIME OF TEST: 04/06/2016 19:51:26

Diagnostics Check		OK	19:44
Air Blank		0.000	19:44
Control Test		0.079	19:45
Air Blank		0.000	19:45
Subject Sample	# 1	0.092	19:46
Breath Volume		1.960	19:46
Air Blank		0.000	19:47
Air Blank		0.000	19:49
Subject Sample	# 2	0.090	19:49
Breath Volume		2.132	19:49
Air Blank		0.000	19:50
Control Test		0.078	19:50
Air Blank		0.000	19:51
Diagnostics Check		OK	19:51

Data Download Date/Time: 04/15/2016 13:04

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CUMMINGS,PETER
ARRESTING OFFICER AGENCY: Lantana PD
LAST AGENCY INSPECTION: 03/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A3
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:57

SUBJECT'S LAST NAME: DOBRANSKY

DATE/TIME OF TEST: 04/07/2016 02:30:11

Diagnostics Check		OK	02:23
Air Blank		0.000	02:24
Control Test		0.079	02:24
Air Blank		0.000	02:25
Subject Sample	# 1	0.000	02:25
Breath Volume		3.054	02:25
Air Blank		0.000	02:26
Air Blank		0.000	02:28
Subject Sample	# 2	0.000	02:28
Breath Volume		3.375	02:28
Air Blank		0.000	02:29
Control Test		0.078	02:29
Air Blank		0.000	02:30
Diagnostics Check		OK	02:30

Data Download Date/Time: 04/15/2016 13:04

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: DERBY,JOHN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 03/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A3
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: CADET

DATE/TIME OF TEST: 04/07/2016 05:12:45

Diagnostics Check		OK	04:59	
Air Blank		0.000	05:00	
Control Test		0.080	05:00	
Air Blank		0.000	05:01	
Subject Sample	# 1	0.216	05:04	Volume Not Met
Breath Volume		0.789	05:04	
Air Blank		0.000	05:05	
Air Blank		0.000	05:06	
Subject Sample	# 2	0.226	05:07	
Breath Volume		1.988	05:07	
Air Blank		0.000	05:08	
Air Blank		0.000	05:09	
Subject Sample	# 3	0.235	05:10	
Breath Volume		2.128	05:10	
Air Blank		0.000	05:11	
Control Test		0.078	05:12	
Air Blank		0.000	05:12	
Diagnostics Check		OK	05:12	

Data Download Date/Time: 04/15/2016 13:04

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAIN,JEANETTE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LERNER,G
ARRESTING OFFICER AGENCY: Palm Beach Shores
LAST AGENCY INSPECTION: 03/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A3
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:01

SUBJECT'S LAST NAME: JULIUS

DATE/TIME OF TEST: 04/07/2016 21:30:07

Diagnostics Check		OK	21:26	
Air Blank		0.000	21:27	
Control Test		0.080	21:27	
Air Blank		0.000	21:28	
Subject Sample	# 1		21:28	Subject Test Refused
Breath Volume		0.000	21:28	
Air Blank		0.000	21:29	
Control Test		0.079	21:29	
Air Blank		0.000	21:30	
Diagnostics Check		OK	21:30	

Data Download Date/Time: 04/15/2016 13:04

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARI,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CARBUCCIA,ABIGAIL
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 03/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A3
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: REEVES

DATE/TIME OF TEST: 04/08/2016 01:53:01

Diagnostics Check		OK	01:43
Air Blank		0.000	01:44
Control Test		0.080	01:44
Air Blank		0.000	01:45
Subject Sample	# 1	0.199	01:46
Breath Volume		1.292	01:46
Air Blank		0.000	01:47
Air Blank		0.000	01:48
Subject Sample	# 2	0.208	01:50
Breath Volume		1.691	01:50
Air Blank		0.000	01:52
Control Test		0.076	01:52
Air Blank		0.000	01:52
Diagnostics Check		OK	01:53

Data Download Date/Time: 04/15/2016 13:04

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MERCIER,DANIEL
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 03/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A3
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:03

SUBJECT'S LAST NAME: DORISCA

DATE/TIME OF TEST: 04/09/2016 00:41:28

Diagnostics Check		OK	00:29	
Air Blank		0.000	00:29	
Control Test		0.079	00:30	
Air Blank		0.000	00:30	
Subject Sample	# 1	0.076	00:34	Volume Not Met
Breath Volume		0.542	00:34	
Air Blank		0.000	00:34	
Air Blank		0.000	00:36	
Subject Sample	# 2	0.082	00:39	Volume Not Met
Breath Volume		0.359	00:39	
Air Blank		0.000	00:40	
Control Test		0.077	00:40	
Air Blank		0.000	00:41	
Diagnostics Check		OK	00:41	

Data Download Date/Time: 04/15/2016 13:04

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MERCIER,DANIEL
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 03/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A3
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:03

SUBJECT'S LAST NAME: DORISCA

DATE/TIME OF TEST: 04/09/2016 00:50:59

Diagnostics Check		OK	00:43
Air Blank		0.000	00:43
Control Test		0.077	00:43
Air Blank		0.000	00:44
Subject Sample	# 1	0.077	00:45
Breath Volume		1.226	00:45
Air Blank		0.000	00:46
Air Blank		0.000	00:48
Subject Sample	# 2	0.084	00:49
Breath Volume		2.042	00:49
Air Blank		0.000	00:50
Control Test		0.078	00:50
Air Blank		0.000	00:50
Diagnostics Check		OK	00:50

Data Download Date/Time: 04/15/2016 13:04

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARI,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITE,EDWARD
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 03/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A3
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: MASLATI

DATE/TIME OF TEST: 04/09/2016 04:13:48

Diagnostics Check		OK	04:07
Air Blank		0.000	04:07
Control Test		0.079	04:07
Air Blank		0.000	04:08
Subject Sample	# 1	0.091	04:08
Breath Volume		2.109	04:08
Air Blank		0.000	04:09
Air Blank		0.000	04:11
Subject Sample	# 2	0.092	04:12
Breath Volume		1.789	04:12
Air Blank		0.000	04:12
Control Test		0.079	04:13
Air Blank		0.000	04:13
Diagnostics Check		OK	04:13

Data Download Date/Time: 04/15/2016 13:04

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARI,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CROY,PAUL
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 03/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A3
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 06:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:14

SUBJECT'S LAST NAME: TORRES

DATE/TIME OF TEST: 04/09/2016 07:46:02

Diagnostics Check		OK	07:39
Air Blank		0.000	07:40
Control Test		0.079	07:40
Air Blank		0.000	07:41
Subject Sample	# 1	0.173	07:41
Breath Volume		2.492	07:41
Air Blank		0.000	07:42
Air Blank		0.000	07:43
Subject Sample	# 2	0.162	07:44
Breath Volume		1.621	07:44
Air Blank		0.000	07:45
Control Test		0.078	07:45
Air Blank		0.000	07:45
Diagnostics Check		OK	07:46

Data Download Date/Time: 04/15/2016 13:04

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARI,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SCHNEIDER,JOSHUA
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 03/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A3
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 08:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:25

SUBJECT'S LAST NAME: MARTINEZ PEREZ

DATE/TIME OF TEST: 04/09/2016 08:59:34

Diagnostics Check		OK	08:52
Air Blank		0.000	08:53
Control Test		0.080	08:53
Air Blank		0.000	08:54
Subject Sample	# 1	0.140	08:54
Breath Volume		1.312	08:54
Air Blank		0.000	08:55
Air Blank		0.000	08:57
Subject Sample	# 2	0.141	08:57
Breath Volume		1.414	08:57
Air Blank		0.000	08:58
Control Test		0.078	08:58
Air Blank		0.000	08:59
Diagnostics Check		OK	08:59

Data Download Date/Time: 04/15/2016 13:04

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: TINEO,NELSON
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 03/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A3
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:12

SUBJECT'S LAST NAME: LUGO ROSALES

DATE/TIME OF TEST: 04/10/2016 02:43:59

Diagnostics Check		OK	02:37
Air Blank		0.000	02:37
Control Test		0.080	02:37
Air Blank		0.000	02:38
Subject Sample	# 1	0.215	02:39
Breath Volume		2.531	02:39
Air Blank		0.000	02:40
Air Blank		0.000	02:41
Subject Sample	# 2	0.219	02:42
Breath Volume		2.992	02:42
Air Blank		0.000	02:43
Control Test		0.078	02:43
Air Blank		0.000	02:43
Diagnostics Check		OK	02:43

Data Download Date/Time: 04/15/2016 13:04

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARI,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: DEROGATIS,JOESPH
ARRESTING OFFICER AGENCY: Palm Springs PD
LAST AGENCY INSPECTION: 03/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A3
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:14

SUBJECT'S LAST NAME: DOMINGO ORTIZ

DATE/TIME OF TEST: 04/10/2016 04:02:28

Diagnostics Check		OK	03:55
Air Blank		0.000	03:56
Control Test		0.079	03:56
Air Blank		0.000	03:57
Subject Sample	# 1	0.211	03:57
Breath Volume		2.671	03:57
Air Blank		0.000	03:58
Air Blank		0.000	04:00
Subject Sample	# 2	0.218	04:00
Breath Volume		1.773	04:00
Air Blank		0.000	04:01
Control Test		0.077	04:01
Air Blank		0.000	04:02
Diagnostics Check		OK	04:02

Data Download Date/Time: 04/15/2016 13:04

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAIN,JEANETTE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FREY,JACOB
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 03/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A3
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: HODEZ

DATE/TIME OF TEST: 04/12/2016 00:37:18

Diagnostics Check		OK	00:30
Air Blank		0.000	00:30
Control Test		0.080	00:31
Air Blank		0.000	00:31
Subject Sample	# 1	0.000	00:32
Breath Volume		1.140	00:32
Air Blank		0.000	00:32
Air Blank		0.000	00:34
Subject Sample	# 2	0.000	00:35
Breath Volume		1.187	00:35
Air Blank		0.000	00:36
Control Test		0.079	00:36
Air Blank		0.000	00:37
Diagnostics Check		OK	00:37

Data Download Date/Time: 04/15/2016 13:04

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARI,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MERCIER,DAANIEL
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 03/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A3
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: OLALDE-GONZALEZ

DATE/TIME OF TEST: 04/15/2016 00:57:40

Diagnostics Check		OK	00:50
Air Blank		0.000	00:51
Control Test		0.080	00:51
Air Blank		0.000	00:52
Subject Sample	# 1	0.236	00:52
Breath Volume		2.890	00:52
Air Blank		0.000	00:53
Air Blank		0.000	00:55
Subject Sample	# 2	0.237	00:55
Breath Volume		3.035	00:55
Air Blank		0.000	00:56
Control Test		0.076	00:57
Air Blank		0.000	00:57
Diagnostics Check		OK	00:57

Data Download Date/Time: 04/15/2016 13:04

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARI,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: GEE,JOHN
ARRESTING OFFICER AGENCY: Palm Springs PD
LAST AGENCY INSPECTION: 03/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A3
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: ALEXANDER

DATE/TIME OF TEST: 04/01/2016 00:19:48

Diagnostics Check		OK	00:13
Air Blank		0.000	00:13
Control Test		0.079	00:13
Air Blank		0.000	00:14
Subject Sample	# 1	0.039	00:15
Breath Volume		1.730	00:15
Air Blank		0.000	00:16
Air Blank		0.000	00:17
Subject Sample	# 2	0.035	00:18
Breath Volume		2.507	00:18
Air Blank		0.000	00:18
Control Test		0.079	00:19
Air Blank		0.000	00:19
Diagnostics Check		OK	00:19

Data Download Date/Time: 04/15/2016 11:43

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITE,KEITH
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 03/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A3
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: LEGEZDH

DATE/TIME OF TEST: 04/01/2016 00:45:28

Diagnostics Check		OK	00:38
Air Blank		0.000	00:39
Control Test		0.078	00:39
Air Blank		0.000	00:40
Subject Sample	# 1	0.190	00:40
Breath Volume		3.210	00:40
Air Blank		0.000	00:41
Air Blank		0.000	00:43
Subject Sample	# 2	0.199	00:43
Breath Volume		2.613	00:43
Air Blank		0.000	00:44
Control Test		0.079	00:44
Air Blank		0.000	00:45
Diagnostics Check		OK	00:45

Data Download Date/Time: 04/15/2016 11:43

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DORSEY,PAMELA,F
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: GRIESER,KEITH
ARRESTING OFFICER AGENCY: North Palm Beach
LAST AGENCY INSPECTION: 03/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A3
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 12:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:46

SUBJECT'S LAST NAME: CROSS

DATE/TIME OF TEST: 04/01/2016 13:37:28

Diagnostics Check		OK	13:30
Air Blank		0.000	13:30
Control Test		0.080	13:31
Air Blank		0.000	13:31
Subject Sample	# 1	0.000	13:32
Breath Volume		4.785	13:32
Air Blank		0.000	13:33
Air Blank		0.000	13:35
Subject Sample	# 2	0.000	13:36
Breath Volume		4.058	13:36
Air Blank		0.000	13:36
Control Test		0.080	13:36
Air Blank		0.000	13:37
Diagnostics Check		OK	13:37

Data Download Date/Time: 04/15/2016 11:43

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MCGRIFF,CALVIN
ARRESTING OFFICER AGENCY: Juno Beach PD
LAST AGENCY INSPECTION: 03/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A3
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:07

SUBJECT'S LAST NAME: ROWELL

DATE/TIME OF TEST: 04/01/2016 23:40:06

Diagnostics Check		OK	23:33
Air Blank		0.000	23:33
Control Test		0.081	23:34
Air Blank		0.000	23:34
Subject Sample	# 1	0.153	23:35
Breath Volume		1.390	23:35
Air Blank		0.000	23:36
Air Blank		0.000	23:37
Subject Sample	# 2	0.153	23:38
Breath Volume		1.714	23:38
Air Blank		0.000	23:39
Control Test		0.080	23:39
Air Blank		0.000	23:40
Diagnostics Check		OK	23:40

Data Download Date/Time: 04/15/2016 11:43

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BORROWS,ANDREW
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 03/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A3
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:48

SUBJECT'S LAST NAME: GUEST

DATE/TIME OF TEST: 04/02/2016 22:21:48

Diagnostics Check		OK	22:13
Air Blank		0.000	22:14
Control Test		0.080	22:14
Air Blank		0.000	22:14
Subject Sample	# 1	0.223	22:16
Breath Volume		1.511	22:16
Air Blank		0.000	22:17
Air Blank		0.000	22:19
Subject Sample	# 2	0.234	22:20
Breath Volume		1.683	22:20
Air Blank		0.000	22:20
Control Test		0.079	22:21
Air Blank		0.000	22:21
Diagnostics Check		OK	22:21

Data Download Date/Time: 04/15/2016 11:43

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DORSEY,PAMELA,F
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MORALES,HERNALDO
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 03/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A3
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 16:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:43

SUBJECT'S LAST NAME: PEREZ CARIAS

DATE/TIME OF TEST: 04/03/2016 18:20:26

Diagnostics Check		OK	18:13
Air Blank		0.000	18:13
Control Test		0.080	18:13
Air Blank		0.000	18:14
Subject Sample	# 1	0.153	18:15
Breath Volume		1.394	18:15
Air Blank		0.000	18:16
Air Blank		0.000	18:17
Subject Sample	# 2	0.156	18:18
Breath Volume		2.054	18:18
Air Blank		0.000	18:19
Control Test		0.078	18:19
Air Blank		0.000	18:20
Diagnostics Check		OK	18:20

Data Download Date/Time: 04/15/2016 11:43

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARI,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: HACHIGIAN,DOMINIC
ARRESTING OFFICER AGENCY: North Palm Beach
LAST AGENCY INSPECTION: 03/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A3
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 14:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:35

SUBJECT'S LAST NAME: CAMERON

DATE/TIME OF TEST: 04/06/2016 16:06:15

Diagnostics Check		OK	15:59
Air Blank		0.000	16:00
Control Test		0.080	16:00
Air Blank		0.000	16:00
Subject Sample	# 1	0.000	16:01
Breath Volume		1.617	16:01
Air Blank		0.000	16:02
Air Blank		0.000	16:04
Subject Sample	# 2	0.000	16:04
Breath Volume		2.156	16:04
Air Blank		0.000	16:05
Control Test		0.080	16:05
Air Blank		0.000	16:06
Diagnostics Check		OK	16:06

Data Download Date/Time: 04/15/2016 11:43

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DORSEY,PAMELA,F
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LASSO,JAVIER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 03/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A3
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 17:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:43

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 04/07/2016 19:15:06

Diagnostics Check		OK	19:06
Air Blank		0.000	19:07
Control Test		0.080	19:07
Air Blank		0.000	19:08
Subject Sample	# 1	0.213	19:08
Breath Volume		3.734	19:08
Air Blank		0.000	19:09
Air Blank		0.000	19:11
Subject Sample	# 2	0.202	19:13
Breath Volume		2.578	19:13
Air Blank		0.000	19:14
Control Test		0.080	19:14
Air Blank		0.000	19:15
Diagnostics Check		OK	19:15

Data Download Date/Time: 04/15/2016 11:43

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DORSEY,PAMELA,F
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: DEROGATIS,JOSEPH
ARRESTING OFFICER AGENCY: Palm Springs PD
LAST AGENCY INSPECTION: 03/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A3
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: NAFIS

DATE/TIME OF TEST: 04/09/2016 00:22:05

Diagnostics Check		OK	00:13
Air Blank		0.000	00:14
Control Test		0.079	00:14
Air Blank		0.000	00:15
Subject Sample	# 1	0.212	00:17
Breath Volume		2.722	00:17
Air Blank		0.000	00:17
Air Blank		0.000	00:19
Subject Sample	# 2	0.213	00:20
Breath Volume		2.097	00:20
Air Blank		0.000	00:21
Control Test		0.077	00:21
Air Blank		0.000	00:21
Diagnostics Check		OK	00:22

Data Download Date/Time: 04/15/2016 11:43

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CHIQUITO RODIGUEZ,LENNDY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 03/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A3
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:19

SUBJECT'S LAST NAME: PEREZ DOMINGO

DATE/TIME OF TEST: 04/09/2016 03:51:04

Diagnostics Check		OK	03:44
Air Blank		0.000	03:44
Control Test		0.079	03:45
Air Blank		0.000	03:45
Subject Sample	# 1	0.214	03:46
Breath Volume		1.925	03:46
Air Blank		0.000	03:47
Air Blank		0.000	03:48
Subject Sample	# 2	0.227	03:49
Breath Volume		3.015	03:49
Air Blank		0.000	03:50
Control Test		0.079	03:50
Air Blank		0.000	03:50
Diagnostics Check		OK	03:51

Data Download Date/Time: 04/15/2016 11:43

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DORSEY,PAMELA,F
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MORA,AARON
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 03/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A3
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 11:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:02

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 04/09/2016 13:30:59

Diagnostics Check		OK	13:26	
Air Blank		0.000	13:26	
Control Test		0.079	13:27	
Air Blank		0.000	13:27	
Subject Sample	# 1		13:29	Subject Test Refused
Breath Volume		0.000	13:29	
Air Blank		0.000	13:30	
Control Test		0.080	13:30	
Air Blank		0.000	13:30	
Diagnostics Check		OK	13:30	

Data Download Date/Time: 04/15/2016 11:43

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: COON,MICHAEL
ARRESTING OFFICER AGENCY: Juno Beach PD
LAST AGENCY INSPECTION: 03/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A3
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:38

SUBJECT'S LAST NAME: BLAZEK

DATE/TIME OF TEST: 04/09/2016 23:10:27

Diagnostics Check		OK	23:03
Air Blank		0.000	23:03
Control Test		0.080	23:04
Air Blank		0.000	23:04
Subject Sample	# 1	0.257	23:05
Breath Volume		1.710	23:05
Air Blank		0.000	23:06
Air Blank		0.000	23:08
Subject Sample	# 2	0.262	23:08
Breath Volume		1.656	23:08
Air Blank		0.000	23:09
Control Test		0.079	23:09
Air Blank		0.000	23:10
Diagnostics Check		OK	23:10

Data Download Date/Time: 04/15/2016 11:43

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARI,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CHIQUITO RODRIGUEZ,LENNDY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 03/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A3
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:07
GENDER: Male

OBSERVATION PERIOD BEGAN: 01:49

SUBJECT'S LAST NAME: ALMARAZ HERNANDEZ

DATE/TIME OF TEST: 04/10/2016 02:22:03

Diagnostics Check		OK	02:15
Air Blank		0.000	02:15
Control Test		0.080	02:16
Air Blank		0.000	02:16
Subject Sample	# 1	0.192	02:17
Breath Volume		1.144	02:17
Air Blank		0.000	02:17
Air Blank		0.000	02:19
Subject Sample	# 2	0.192	02:20
Breath Volume		1.812	02:20
Air Blank		0.000	02:21
Control Test		0.079	02:21
Air Blank		0.000	02:21
Diagnostics Check		OK	02:22

Data Download Date/Time: 04/15/2016 11:43

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARI,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MERCIER,DANIEL
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 03/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A3
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:43

SUBJECT'S LAST NAME: WOO

DATE/TIME OF TEST: 04/10/2016 03:16:43

Diagnostics Check		OK	03:09
Air Blank		0.000	03:09
Control Test		0.080	03:10
Air Blank		0.000	03:10
Subject Sample	# 1	0.228	03:11
Breath Volume		1.222	03:11
Air Blank		0.000	03:12
Air Blank		0.000	03:13
Subject Sample	# 2	0.229	03:15
Breath Volume		1.164	03:15
Air Blank		0.000	03:15
Control Test		0.079	03:16
Air Blank		0.000	03:16
Diagnostics Check		OK	03:16

Data Download Date/Time: 04/15/2016 11:43

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DORSEY,PAMELA,F
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: JOHNSON,EDWIN
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 03/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A3
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 09:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:23

SUBJECT'S LAST NAME: BORRELLI

DATE/TIME OF TEST: 04/10/2016 11:53:30

Diagnostics Check		OK	11:46
Air Blank		0.000	11:46
Control Test		0.080	11:47
Air Blank		0.000	11:47
Subject Sample	# 1	0.000	11:48
Breath Volume		2.386	11:48
Air Blank		0.000	11:49
Air Blank		0.000	11:51
Subject Sample	# 2	0.000	11:52
Breath Volume		2.933	11:52
Air Blank		0.000	11:52
Control Test		0.080	11:52
Air Blank		0.000	11:53
Diagnostics Check		OK	11:53

Data Download Date/Time: 04/15/2016 11:43

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MIRANDA,EDWIN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 03/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A3
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: CHAVAJAY PEREZ

DATE/TIME OF TEST: 04/12/2016 01:12:12

Diagnostics Check		OK	01:05
Air Blank		0.000	01:05
Control Test		0.080	01:05
Air Blank		0.000	01:06
Subject Sample	# 1	0.131	01:07
Breath Volume		1.781	01:07
Air Blank		0.000	01:07
Air Blank		0.000	01:09
Subject Sample	# 2	0.132	01:10
Breath Volume		1.972	01:10
Air Blank		0.000	01:11
Control Test		0.080	01:11
Air Blank		0.000	01:12
Diagnostics Check		OK	01:12

Data Download Date/Time: 04/15/2016 11:43

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FANDREY,CHRISTOPHER
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 03/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A3
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:58

SUBJECT'S LAST NAME: SANDOVAL SANTOS

DATE/TIME OF TEST: 04/13/2016 01:35:39

Diagnostics Check		OK	01:23
Air Blank		0.000	01:24
Control Test		0.079	01:24
Air Blank		0.000	01:24
Subject Sample	# 1	0.196	01:26
Breath Volume		1.367	01:26
Air Blank		0.000	01:27
Air Blank		0.000	01:28
Subject Sample	# 2	0.227	01:29
Breath Volume		1.679	01:29
Air Blank		0.000	01:30
Air Blank		0.000	01:32
Subject Sample	# 3	0.202	01:33
Breath Volume		1.875	01:33
Air Blank		0.000	01:34
Control Test		0.078	01:35
Air Blank		0.000	01:35
Diagnostics Check		OK	01:35

Data Download Date/Time: 04/15/2016 11:43

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: RODRIGUEZ,JOSE
ARRESTING OFFICER AGENCY: Palm Beach PD
LAST AGENCY INSPECTION: 03/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A3
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: SCHALLER

DATE/TIME OF TEST: 04/14/2016 04:06:05

Diagnostics Check		OK	04:03	
Air Blank		0.000	04:03	
Control Test		0.080	04:04	
Air Blank		0.000	04:04	
Subject Sample	# 1	0.177	04:05	
Breath Volume		3.925	04:05	
Air Blank			04:05	RFI Detect
Air Blank			04:05	RFI Detect
Air Blank		0.000	04:06	

Data Download Date/Time: 04/15/2016 11:43

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: RODRIGUEZ,JOSE
ARRESTING OFFICER AGENCY: Palm Beach PD
LAST AGENCY INSPECTION: 03/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A3
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: SCHALLER

DATE/TIME OF TEST: 04/14/2016 04:13:42

Diagnostics Check		OK	04:07
Air Blank		0.000	04:07
Control Test		0.079	04:08
Air Blank		0.000	04:08
Subject Sample	# 1	0.180	04:09
Breath Volume		4.429	04:09
Air Blank		0.000	04:09
Air Blank		0.000	04:11
Subject Sample	# 2	0.171	04:12
Breath Volume		3.683	04:12
Air Blank		0.000	04:12
Control Test		0.079	04:13
Air Blank		0.000	04:13
Diagnostics Check		OK	04:13

Data Download Date/Time: 04/15/2016 11:43

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DORSEY,PAMELA,F
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: GEE,JOHN
ARRESTING OFFICER AGENCY: Palm Springs PD
LAST AGENCY INSPECTION: 03/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A3
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:07

SUBJECT'S LAST NAME: OGDEN

DATE/TIME OF TEST: 04/14/2016 21:41:16

Diagnostics Check		OK	21:33
Air Blank		0.000	21:33
Control Test		0.081	21:34
Air Blank		0.000	21:34
Subject Sample	# 1	0.034	21:36
Breath Volume		1.531	21:36
Air Blank		0.000	21:37
Air Blank		0.000	21:39
Subject Sample	# 2	0.033	21:39
Breath Volume		1.246	21:39
Air Blank		0.000	21:40
Control Test		0.080	21:40
Air Blank		0.000	21:41
Diagnostics Check		OK	21:41

Data Download Date/Time: 04/15/2016 11:43

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: PUJOL,JESUS
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 03/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A3
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:48

SUBJECT'S LAST NAME: COTE

DATE/TIME OF TEST: 04/15/2016 00:21:43

Diagnostics Check		OK	00:15
Air Blank		0.000	00:15
Control Test		0.079	00:15
Air Blank		0.000	00:16
Subject Sample	# 1	0.216	00:17
Breath Volume		2.546	00:17
Air Blank		0.000	00:17
Air Blank		0.000	00:19
Subject Sample	# 2	0.221	00:20
Breath Volume		2.832	00:20
Air Blank		0.000	00:20
Control Test		0.079	00:21
Air Blank		0.000	00:21
Diagnostics Check		OK	00:21

Data Download Date/Time: 04/15/2016 11:43

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FERANDEZ,RYAN
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 04/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 01:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: BATES

DATE/TIME OF TEST: 04/16/2016 02:25:16

Diagnostics Check		OK	02:09	
Air Blank		0.000	02:09	
Control Test		0.081	02:10	
Air Blank		0.000	02:10	
Subject Sample	# 1	0.154	02:13	Volume Not Met
Breath Volume		0.117	02:13	
Air Blank		0.000	02:14	
Air Blank		0.000	02:16	
Subject Sample	# 2	0.173	02:18	
Breath Volume		1.390	02:18	
Air Blank		0.000	02:19	
Air Blank		0.000	02:20	
Subject Sample	# 3	0.183	02:23	
Breath Volume		2.117	02:23	
Air Blank		0.000	02:24	
Control Test		0.080	02:24	
Air Blank		0.000	02:25	
Diagnostics Check		OK	02:25	

Data Download Date/Time: 05/20/2016 14:00

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHAFER,NANCY,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITE,E
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 04/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 19:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:56

SUBJECT'S LAST NAME: CABERA

DATE/TIME OF TEST: 04/16/2016 20:39:38

Diagnostics Check		OK	20:32
Air Blank		0.000	20:32
Control Test		0.081	20:33
Air Blank		0.000	20:33
Subject Sample	# 1	0.241	20:34
Breath Volume		1.785	20:34
Air Blank		0.000	20:35
Air Blank		0.000	20:37
Subject Sample	# 2	0.234	20:37
Breath Volume		2.652	20:37
Air Blank		0.000	20:38
Control Test		0.080	20:39
Air Blank		0.000	20:39
Diagnostics Check		OK	20:39

Data Download Date/Time: 05/20/2016 14:00

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FERNANDEZ,RYAN
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 04/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 00:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:08

SUBJECT'S LAST NAME: ZANDERS III

DATE/TIME OF TEST: 04/17/2016 01:43:13

Diagnostics Check		OK	01:35
Air Blank		0.000	01:36
Control Test		0.081	01:36
Air Blank		0.000	01:36
Subject Sample	# 1	0.155	01:38
Breath Volume		1.281	01:38
Air Blank		0.000	01:39
Air Blank		0.000	01:40
Subject Sample	# 2	0.158	01:41
Breath Volume		1.304	01:41
Air Blank		0.000	01:42
Control Test		0.080	01:42
Air Blank		0.000	01:43
Diagnostics Check		OK	01:43

Data Download Date/Time: 05/20/2016 14:00

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BUSH,SHANA
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 04/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 06:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:33

SUBJECT'S LAST NAME: CORONA JR

DATE/TIME OF TEST: 04/17/2016 08:07:12

Diagnostics Check		OK	08:00
Air Blank		0.000	08:00
Control Test		0.082	08:01
Air Blank		0.000	08:01
Subject Sample	# 1	0.138	08:02
Breath Volume		1.894	08:02
Air Blank		0.000	08:03
Air Blank		0.000	08:05
Subject Sample	# 2	0.139	08:05
Breath Volume		1.613	08:05
Air Blank		0.000	08:06
Control Test		0.081	08:06
Air Blank		0.000	08:07
Diagnostics Check		OK	08:07

Data Download Date/Time: 05/20/2016 14:00

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITE,EDWARD
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 04/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 20:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:47

SUBJECT'S LAST NAME: LEMONS

DATE/TIME OF TEST: 04/18/2016 21:20:19

Diagnostics Check		OK	21:13
Air Blank		0.000	21:14
Control Test		0.082	21:14
Air Blank		0.000	21:14
Subject Sample	# 1	0.000	21:15
Breath Volume		1.394	21:15
Air Blank		0.000	21:16
Air Blank		0.000	21:18
Subject Sample	# 2	0.000	21:18
Breath Volume		1.496	21:18
Air Blank		0.000	21:19
Control Test		0.081	21:19
Air Blank		0.000	21:20
Diagnostics Check		OK	21:20

Data Download Date/Time: 05/20/2016 14:00

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MORAL,LONNEY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 04/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 21:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:35

SUBJECT'S LAST NAME: STITES

DATE/TIME OF TEST: 04/18/2016 22:21:52

Diagnostics Check		OK	22:15
Air Blank		0.000	22:15
Control Test		0.082	22:15
Air Blank		0.000	22:16
Subject Sample	# 1	0.143	22:17
Breath Volume		1.703	22:17
Air Blank		0.000	22:17
Air Blank		0.000	22:19
Subject Sample	# 2	0.143	22:20
Breath Volume		1.320	22:20
Air Blank		0.000	22:20
Control Test		0.082	22:21
Air Blank		0.000	22:21
Diagnostics Check		OK	22:21

Data Download Date/Time: 05/20/2016 14:00

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAIN,JEANETTE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FREY,JACOB
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 04/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 22:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: MONI

DATE/TIME OF TEST: 04/20/2016 00:37:59

Diagnostics Check		OK	00:23	
Air Blank		0.000	00:23	
Control Test		0.082	00:24	
Air Blank		0.000	00:24	
Subject Sample	# 1	0.000	00:25	
Breath Volume		1.652	00:25	
Air Blank		0.000	00:26	
Air Blank		0.000	00:28	
Subject Sample	# 2	0.000	00:31	Volume Not Met
Breath Volume		0.421	00:31	
Air Blank		0.000	00:32	
Air Blank		0.000	00:34	
Subject Sample	# 3	0.000	00:36	
Breath Volume		1.261	00:36	
Air Blank		0.000	00:37	
Control Test		0.081	00:37	
Air Blank		0.000	00:37	
Diagnostics Check		OK	00:37	

Data Download Date/Time: 05/20/2016 14:00

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MERCIER,DANIEL
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 04/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 01:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:28

SUBJECT'S LAST NAME: MURILLO SANCHEZ

DATE/TIME OF TEST: 04/22/2016 02:01:33

Diagnostics Check		OK	01:52
Air Blank		0.000	01:52
Control Test		0.081	01:53
Air Blank		0.000	01:53
Subject Sample	# 1	0.179	01:56
Breath Volume		1.207	01:56
Air Blank		0.000	01:57
Air Blank		0.000	01:59
Subject Sample	# 2	0.184	01:59
Breath Volume		1.269	01:59
Air Blank		0.000	02:00
Control Test		0.079	02:00
Air Blank		0.000	02:01
Diagnostics Check		OK	02:01

Data Download Date/Time: 05/20/2016 14:00

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARI,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SCHAEFER,JOHN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 04/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 04:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:00

SUBJECT'S LAST NAME: CASTILLO

DATE/TIME OF TEST: 04/22/2016 05:34:40

Diagnostics Check		OK	05:28
Air Blank		0.000	05:28
Control Test		0.081	05:29
Air Blank		0.000	05:29
Subject Sample	# 1	0.227	05:30
Breath Volume		1.914	05:30
Air Blank		0.000	05:30
Air Blank		0.000	05:32
Subject Sample	# 2	0.229	05:33
Breath Volume		1.652	05:33
Air Blank		0.000	05:33
Control Test		0.081	05:34
Air Blank		0.000	05:34
Diagnostics Check		OK	05:34

Data Download Date/Time: 05/20/2016 14:00

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARI,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: GEE,JOHN
ARRESTING OFFICER AGENCY: Palm Springs PD
LAST AGENCY INSPECTION: 04/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 01:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:39

SUBJECT'S LAST NAME: PEDRO JULIAN

DATE/TIME OF TEST: 04/23/2016 02:13:58

Diagnostics Check		OK	02:07
Air Blank		0.000	02:07
Control Test		0.081	02:07
Air Blank		0.000	02:08
Subject Sample	# 1	0.278	02:09
Breath Volume		1.148	02:09
Air Blank		0.000	02:09
Air Blank		0.000	02:11
Subject Sample	# 2	0.280	02:12
Breath Volume		1.753	02:12
Air Blank		0.000	02:13
Control Test		0.080	02:13
Air Blank		0.000	02:13
Diagnostics Check		OK	02:13

Data Download Date/Time: 05/20/2016 14:00

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARI,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FEQUIERE,RALPH
ARRESTING OFFICER AGENCY: Palm Springs PD
LAST AGENCY INSPECTION: 04/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 04:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:50

SUBJECT'S LAST NAME: MENDEZ

DATE/TIME OF TEST: 04/23/2016 05:25:25

Diagnostics Check		OK	05:18
Air Blank		0.000	05:18
Control Test		0.082	05:19
Air Blank		0.000	05:19
Subject Sample	# 1	0.082	05:20
Breath Volume		1.234	05:20
Air Blank		0.000	05:21
Air Blank		0.000	05:23
Subject Sample	# 2	0.084	05:23
Breath Volume		1.464	05:23
Air Blank		0.000	05:24
Control Test		0.081	05:24
Air Blank		0.000	05:25
Diagnostics Check		OK	05:25

Data Download Date/Time: 05/20/2016 14:00

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BLACKMAN,VICTOR
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 04/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 03:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:02

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 04/24/2016 04:49:39

Diagnostics Check		OK	04:42
Air Blank		0.000	04:42
Control Test		0.082	04:43
Air Blank		0.000	04:43
Subject Sample	# 1	0.192	04:45
Breath Volume		1.402	04:45
Air Blank		0.000	04:45
Air Blank		0.000	04:47
Subject Sample	# 2	0.187	04:47
Breath Volume		1.714	04:47
Air Blank		0.000	04:48
Control Test		0.079	04:49
Air Blank		0.000	04:49
Diagnostics Check		OK	04:49

Data Download Date/Time: 05/20/2016 14:00

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WARD,CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 04/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 04:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:02

SUBJECT'S LAST NAME: MIOZZI

DATE/TIME OF TEST: 04/24/2016 06:41:15

Diagnostics Check		OK	06:30
Air Blank		0.000	06:30
Control Test		0.082	06:31
Air Blank		0.000	06:31
Subject Sample	# 1	0.165	06:33
Breath Volume		1.226	06:33
Air Blank		0.000	06:34
Air Blank		0.000	06:36
Subject Sample	# 2	0.171	06:39
Breath Volume		4.265	06:39
Air Blank		0.000	06:40
Control Test		0.080	06:40
Air Blank		0.000	06:41
Diagnostics Check		OK	06:41

Data Download Date/Time: 05/20/2016 14:00

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAIN,JEANETTE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SCHAEFER,JOHN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 04/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 21:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: SAUNDERS

DATE/TIME OF TEST: 04/24/2016 22:02:18

Diagnostics Check		OK	21:55
Air Blank		0.000	21:55
Control Test		0.081	21:56
Air Blank		0.000	21:56
Subject Sample	# 1	0.142	21:57
Breath Volume		1.152	21:57
Air Blank		0.000	21:58
Air Blank		0.000	22:00
Subject Sample	# 2	0.154	22:00
Breath Volume		2.140	22:00
Air Blank		0.000	22:01
Control Test		0.081	22:01
Air Blank		0.000	22:02
Diagnostics Check		OK	22:02

Data Download Date/Time: 05/20/2016 14:00

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: GARCIA,ISRAEL
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 04/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 01:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:28

SUBJECT'S LAST NAME: LOPEZ PEREZ

DATE/TIME OF TEST: 04/25/2016 03:24:30

Diagnostics Check		OK	03:16
Air Blank		0.000	03:17
Control Test		0.081	03:17
Air Blank		0.000	03:17
Subject Sample	# 1	0.201	03:19
Breath Volume		1.183	03:19
Air Blank		0.000	03:20
Air Blank		0.000	03:22
Subject Sample	# 2	0.192	03:22
Breath Volume		1.285	03:22
Air Blank		0.000	03:23
Control Test		0.080	03:23
Air Blank		0.000	03:24
Diagnostics Check		OK	03:24

Data Download Date/Time: 05/20/2016 14:00

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAIN,JEANETTE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MILLER,STEPHEN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 04/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 21:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: VERTULLO

DATE/TIME OF TEST: 04/26/2016 22:55:33

Diagnostics Check		OK	22:48
Air Blank		0.000	22:48
Control Test		0.082	22:49
Air Blank		0.000	22:49
Subject Sample	# 1	0.092	22:51
Breath Volume		1.714	22:51
Air Blank		0.000	22:51
Air Blank		0.000	22:53
Subject Sample	# 2	0.092	22:53
Breath Volume		1.980	22:53
Air Blank		0.000	22:54
Control Test		0.082	22:54
Air Blank		0.000	22:55
Diagnostics Check		OK	22:55

Data Download Date/Time: 05/20/2016 14:00

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARI,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: KWAKU,EDWARD
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 04/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 23:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:28

SUBJECT'S LAST NAME: MESHEL

DATE/TIME OF TEST: 04/28/2016 01:00:42

Diagnostics Check		OK	00:54
Air Blank		0.000	00:55
Control Test		0.082	00:55
Air Blank		0.000	00:55
Subject Sample	# 1	0.156	00:56
Breath Volume		1.773	00:56
Air Blank		0.000	00:56
Air Blank		0.000	00:58
Subject Sample	# 2	0.155	00:59
Breath Volume		1.511	00:59
Air Blank		0.000	00:59
Control Test		0.081	01:00
Air Blank		0.000	01:00
Diagnostics Check		OK	01:00

Data Download Date/Time: 05/20/2016 14:00

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: GODDEN,JOHN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 04/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 08:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:50

SUBJECT'S LAST NAME: OLOUGHLIN

DATE/TIME OF TEST: 04/28/2016 11:45:22

Diagnostics Check		OK	11:38
Air Blank		0.000	11:39
Control Test		0.082	11:39
Air Blank		0.000	11:39
Subject Sample	# 1	0.000	11:40
Breath Volume		3.878	11:40
Air Blank		0.000	11:41
Air Blank		0.000	11:43
Subject Sample	# 2	0.000	11:43
Breath Volume		2.515	11:43
Air Blank		0.000	11:44
Control Test		0.081	11:44
Air Blank		0.000	11:45
Diagnostics Check		OK	11:45

Data Download Date/Time: 05/20/2016 14:00

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ANDERSON,O'NEIL
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 04/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 00:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:32

SUBJECT'S LAST NAME: CASTRO-OSORIO

DATE/TIME OF TEST: 04/29/2016 04:11:25

Diagnostics Check		OK	04:04
Air Blank		0.000	04:05
Control Test		0.082	04:05
Air Blank		0.000	04:05
Subject Sample	# 1	0.144	04:06
Breath Volume		2.937	04:06
Air Blank		0.000	04:07
Air Blank		0.000	04:09
Subject Sample	# 2	0.144	04:09
Breath Volume		2.718	04:09
Air Blank		0.000	04:10
Control Test		0.081	04:10
Air Blank		0.000	04:11
Diagnostics Check		OK	04:11

Data Download Date/Time: 05/20/2016 14:00

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MERCIER,DANIEL
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 04/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 22:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: TEVALAN DELEON

DATE/TIME OF TEST: 04/30/2016 00:07:49

Diagnostics Check		OK	23:59
Air Blank		0.000	23:59
Control Test		0.081	00:00
Air Blank		0.000	00:00
Subject Sample	# 1	0.150	00:02
Breath Volume		1.843	00:02
Air Blank		0.000	00:03
Air Blank		0.000	00:04
Subject Sample	# 2	0.139	00:06
Breath Volume		1.492	00:06
Air Blank		0.000	00:06
Control Test		0.081	00:07
Air Blank		0.000	00:07
Diagnostics Check		OK	00:07

Data Download Date/Time: 05/20/2016 14:00

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: DOOLEY,ROBERT
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 04/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 14:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:31

SUBJECT'S LAST NAME: MEZA

DATE/TIME OF TEST: 04/30/2016 15:47:57

Diagnostics Check		OK	15:41
Air Blank		0.000	15:41
Control Test		0.082	15:42
Air Blank		0.000	15:42
Subject Sample	# 1	0.173	15:43
Breath Volume		2.535	15:43
Air Blank		0.000	15:44
Air Blank		0.000	15:45
Subject Sample	# 2	0.180	15:46
Breath Volume		2.363	15:46
Air Blank		0.000	15:47
Control Test		0.081	15:47
Air Blank		0.000	15:47
Diagnostics Check		OK	15:47

Data Download Date/Time: 05/20/2016 14:00

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DORSEY,PAMELA,F
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: HANTON,MELINDA
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 04/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A3
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:31

SUBJECT'S LAST NAME: CARTIER

DATE/TIME OF TEST: 04/15/2016 22:03:57

Diagnostics Check		OK	21:56
Air Blank		0.000	21:56
Control Test		0.080	21:57
Air Blank		0.000	21:57
Subject Sample	# 1	0.126	21:58
Breath Volume		2.082	21:58
Air Blank		0.000	21:59
Air Blank		0.000	22:01
Subject Sample	# 2	0.122	22:02
Breath Volume		3.085	22:02
Air Blank		0.000	22:02
Control Test		0.079	22:03
Air Blank		0.000	22:03
Diagnostics Check		OK	22:03

Data Download Date/Time: 05/20/2016 12:22

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WALTON,THOMAS
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 04/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A3
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: BILLECI

DATE/TIME OF TEST: 04/17/2016 00:25:05

Diagnostics Check		OK	00:18
Air Blank		0.000	00:19
Control Test		0.079	00:19
Air Blank		0.000	00:19
Subject Sample	# 1	0.169	00:20
Breath Volume		1.378	00:20
Air Blank		0.000	00:21
Air Blank		0.000	00:22
Subject Sample	# 2	0.169	00:23
Breath Volume		2.894	00:23
Air Blank		0.000	00:24
Control Test		0.079	00:24
Air Blank		0.000	00:25
Diagnostics Check		OK	00:25

Data Download Date/Time: 05/20/2016 12:22

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: HANTON,MELINDA
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 04/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A3
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: UNGER

DATE/TIME OF TEST: 04/17/2016 02:30:59

Diagnostics Check		OK	02:19	
Air Blank		0.000	02:19	
Control Test		0.079	02:19	
Air Blank		0.000	02:20	
Subject Sample	# 1	0.134	02:23	Volume Not Met
Breath Volume		0.039	02:23	
Air Blank		0.000	02:24	
Air Blank		0.000	02:26	
Subject Sample	# 2	0.111	02:29	Volume Not Met
Breath Volume		0.054	02:29	
Air Blank		0.000	02:30	
Control Test		0.077	02:30	
Air Blank		0.000	02:30	
Diagnostics Check		OK	02:30	

Data Download Date/Time: 05/20/2016 12:22

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,5
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: HANTON,MELINDA
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 04/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A3
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: UNGER

DATE/TIME OF TEST: 04/17/2016 02:39:45

Diagnostics Check		OK	02:32
Air Blank		0.000	02:32
Control Test		0.077	02:33
Air Blank		0.000	02:33
Subject Sample	# 1	0.147	02:34
Breath Volume		1.726	02:34
Air Blank		0.000	02:35
Air Blank		0.000	02:37
Subject Sample	# 2	0.140	02:38
Breath Volume		1.164	02:38
Air Blank		0.000	02:38
Control Test		0.077	02:39
Air Blank		0.000	02:39
Diagnostics Check		OK	02:39

Data Download Date/Time: 05/20/2016 12:22

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SAYIN,ROBERT
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 04/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A3
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:28

SUBJECT'S LAST NAME: PENDLETON

DATE/TIME OF TEST: 04/18/2016 00:58:36

Diagnostics Check		OK	00:51
Air Blank		0.000	00:52
Control Test		0.080	00:52
Air Blank		0.000	00:52
Subject Sample	# 1	0.076	00:53
Breath Volume		2.574	00:53
Air Blank		0.000	00:54
Air Blank		0.000	00:56
Subject Sample	# 2	0.078	00:56
Breath Volume		2.363	00:56
Air Blank		0.000	00:57
Control Test		0.079	00:57
Air Blank		0.000	00:58
Diagnostics Check		OK	00:58

Data Download Date/Time: 05/20/2016 12:22

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARI,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITE,EDWARD
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 04/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A3
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:24

SUBJECT'S LAST NAME: MAHMUD

DATE/TIME OF TEST: 04/19/2016 22:56:49

Diagnostics Check		OK	22:50
Air Blank		0.000	22:50
Control Test		0.080	22:51
Air Blank		0.000	22:51
Subject Sample	# 1	0.127	22:52
Breath Volume		1.800	22:52
Air Blank		0.000	22:53
Air Blank		0.000	22:54
Subject Sample	# 2	0.109	22:55
Breath Volume		1.757	22:55
Air Blank		0.000	22:55
Control Test		0.079	22:56
Air Blank		0.000	22:56
Diagnostics Check		OK	22:56

Data Download Date/Time: 05/20/2016 12:22

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CUMMINGS,PETER
ARRESTING OFFICER AGENCY: Lantana PD
LAST AGENCY INSPECTION: 04/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A3
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: SINGH

DATE/TIME OF TEST: 04/22/2016 02:09:38

Diagnostics Check		OK	02:03
Air Blank		0.000	02:03
Control Test		0.079	02:03
Air Blank		0.000	02:04
Subject Sample	# 1	0.186	02:05
Breath Volume		2.523	02:05
Air Blank		0.000	02:05
Air Blank		0.000	02:07
Subject Sample	# 2	0.190	02:07
Breath Volume		2.488	02:07
Air Blank		0.000	02:08
Control Test		0.078	02:08
Air Blank		0.000	02:09
Diagnostics Check		OK	02:09

Data Download Date/Time: 05/20/2016 12:22

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: GEE,JOHN
ARRESTING OFFICER AGENCY: Palm Springs PD
LAST AGENCY INSPECTION: 04/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A3
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: TRUJILLO

DATE/TIME OF TEST: 04/24/2016 00:09:06

Diagnostics Check		OK	00:01
Air Blank		0.000	00:02
Control Test		0.081	00:02
Air Blank		0.000	00:03
Subject Sample	# 1	0.164	00:04
Breath Volume		1.613	00:04
Air Blank		0.000	00:04
Air Blank		0.000	00:06
Subject Sample	# 2	0.169	00:07
Breath Volume		1.429	00:07
Air Blank		0.000	00:08
Control Test		0.080	00:08
Air Blank		0.000	00:09
Diagnostics Check		OK	00:09

Data Download Date/Time: 05/20/2016 12:22

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITE,KEITH
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 04/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A3
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:02

SUBJECT'S LAST NAME: CONTRERAS CALDERON

DATE/TIME OF TEST: 04/24/2016 04:32:17

Diagnostics Check		OK	04:25
Air Blank		0.000	04:25
Control Test		0.080	04:26
Air Blank		0.000	04:26
Subject Sample	# 1	0.178	04:27
Breath Volume		1.980	04:27
Air Blank		0.000	04:28
Air Blank		0.000	04:30
Subject Sample	# 2	0.179	04:30
Breath Volume		2.593	04:30
Air Blank		0.000	04:31
Control Test		0.078	04:31
Air Blank		0.000	04:32
Diagnostics Check		OK	04:32

Data Download Date/Time: 05/20/2016 12:22

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WALTON,THOMAS
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 04/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A3
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:22

SUBJECT'S LAST NAME: SEAHA

DATE/TIME OF TEST: 04/25/2016 02:54:30

Diagnostics Check		OK	02:47
Air Blank		0.000	02:47
Control Test		0.080	02:48
Air Blank		0.000	02:48
Subject Sample	# 1	0.290	02:49
Breath Volume		3.832	02:49
Air Blank		0.000	02:50
Air Blank		0.000	02:52
Subject Sample	# 2	0.287	02:52
Breath Volume		4.585	02:52
Air Blank		0.000	02:53
Control Test		0.078	02:53
Air Blank		0.000	02:54
Diagnostics Check		OK	02:54

Data Download Date/Time: 05/20/2016 12:22

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITE,KEITH
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 04/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A3
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:18

SUBJECT'S LAST NAME: ROY FEQUIERE

DATE/TIME OF TEST: 04/25/2016 03:48:04

Diagnostics Check		OK	03:41
Air Blank		0.000	03:41
Control Test		0.079	03:41
Air Blank		0.000	03:42
Subject Sample	# 1	0.111	03:43
Breath Volume		3.476	03:43
Air Blank		0.000	03:44
Air Blank		0.000	03:45
Subject Sample	# 2	0.111	03:46
Breath Volume		3.292	03:46
Air Blank		0.000	03:47
Control Test		0.079	03:47
Air Blank		0.000	03:47
Diagnostics Check		OK	03:48

Data Download Date/Time: 05/20/2016 12:22

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: HILL,JEREMY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 04/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A3
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: CRUZ-CERVANTES

DATE/TIME OF TEST: 04/26/2016 03:31:54

Diagnostics Check		OK	03:25
Air Blank		0.000	03:25
Control Test		0.081	03:26
Air Blank		0.000	03:26
Subject Sample	# 1	0.162	03:27
Breath Volume		2.671	03:27
Air Blank		0.000	03:28
Air Blank		0.000	03:29
Subject Sample	# 2	0.155	03:30
Breath Volume		1.968	03:30
Air Blank		0.000	03:30
Control Test		0.080	03:31
Air Blank		0.000	03:31
Diagnostics Check		OK	03:31

Data Download Date/Time: 05/20/2016 12:22

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITE,EDWARD
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 04/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A3
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: WOODS

DATE/TIME OF TEST: 04/26/2016 04:08:44

Diagnostics Check		OK	04:02
Air Blank		0.000	04:03
Control Test		0.080	04:03
Air Blank		0.000	04:03
Subject Sample	# 1	0.106	04:04
Breath Volume		1.886	04:04
Air Blank		0.000	04:04
Air Blank		0.000	04:06
Subject Sample	# 2	0.104	04:07
Breath Volume		1.792	04:07
Air Blank		0.000	04:07
Control Test		0.080	04:08
Air Blank		0.000	04:08
Diagnostics Check		OK	04:08

Data Download Date/Time: 05/20/2016 12:22

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: POINTU,PATRICK
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 04/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A3
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: HAGOOD

DATE/TIME OF TEST: 04/27/2016 03:20:35

Diagnostics Check		OK	03:14
Air Blank		0.000	03:14
Control Test		0.080	03:15
Air Blank		0.000	03:15
Subject Sample	# 1	0.149	03:15
Breath Volume		3.203	03:15
Air Blank		0.000	03:16
Air Blank		0.000	03:18
Subject Sample	# 2	0.145	03:18
Breath Volume		3.195	03:18
Air Blank		0.000	03:19
Control Test		0.079	03:19
Air Blank		0.000	03:20
Diagnostics Check		OK	03:20

Data Download Date/Time: 05/20/2016 12:22

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DORSEY,PAMELA,F
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BORROWS,ANDREW
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 04/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A3
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:55

SUBJECT'S LAST NAME: CUTLIP

DATE/TIME OF TEST: 04/27/2016 22:26:49

Diagnostics Check		OK	22:19
Air Blank		0.000	22:20
Control Test		0.081	22:20
Air Blank		0.000	22:20
Subject Sample	# 1	0.130	22:21
Breath Volume		2.050	22:21
Air Blank		0.000	22:22
Air Blank		0.000	22:24
Subject Sample	# 2	0.132	22:25
Breath Volume		2.804	22:25
Air Blank		0.000	22:25
Control Test		0.080	22:26
Air Blank		0.000	22:26
Diagnostics Check		OK	22:26

Data Download Date/Time: 05/20/2016 12:22

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DORSEY,PAMELA,F
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SCHNEIDER,JOSHUA
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 04/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A3
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 14:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:40

SUBJECT'S LAST NAME: KEYS

DATE/TIME OF TEST: 04/28/2016 15:53:23

Diagnostics Check		OK	15:46
Air Blank		0.000	15:46
Control Test		0.080	15:46
Air Blank		0.000	15:47
Subject Sample	# 1	0.000	15:48
Breath Volume		2.906	15:48
Air Blank		0.000	15:49
Air Blank		0.000	15:51
Subject Sample	# 2	0.000	15:51
Breath Volume		2.890	15:51
Air Blank		0.000	15:52
Control Test		0.080	15:52
Air Blank		0.000	15:53
Diagnostics Check		OK	15:53

Data Download Date/Time: 05/20/2016 12:22

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARI,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: PUJOL,JESUS
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 04/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A3
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:59

SUBJECT'S LAST NAME: ABRAMSON

DATE/TIME OF TEST: 04/29/2016 23:33:58

Diagnostics Check		OK	23:27
Air Blank		0.000	23:27
Control Test		0.081	23:28
Air Blank		0.000	23:28
Subject Sample	# 1	0.157	23:29
Breath Volume		2.625	23:29
Air Blank		0.000	23:30
Air Blank		0.000	23:31
Subject Sample	# 2	0.138	23:32
Breath Volume		1.121	23:32
Air Blank		0.000	23:33
Control Test		0.079	23:33
Air Blank		0.000	23:33
Diagnostics Check		OK	23:33

Data Download Date/Time: 05/20/2016 12:22

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITE,KEITH
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 04/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A3
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:24

SUBJECT'S LAST NAME: KEYES

DATE/TIME OF TEST: 04/30/2016 23:53:51

Diagnostics Check		OK	23:47
Air Blank		0.000	23:47
Control Test		0.080	23:48
Air Blank		0.000	23:48
Subject Sample	# 1	0.146	23:49
Breath Volume		2.550	23:49
Air Blank		0.000	23:50
Air Blank		0.000	23:51
Subject Sample	# 2	0.144	23:52
Breath Volume		2.781	23:52
Air Blank		0.000	23:52
Control Test		0.079	23:53
Air Blank		0.000	23:53
Diagnostics Check		OK	23:53

Data Download Date/Time: 05/20/2016 12:22

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARTIGAS,KRISTINA,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: ARTIGAS,KRISTINA
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 03/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 00:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:34

SUBJECT'S LAST NAME: SANCHEZ GOMEZ

DATE/TIME OF TEST: 04/01/2016 02:20:41

Diagnostics Check		OK	02:13
Air Blank		0.000	02:14
Control Test		0.079	02:14
Air Blank		0.000	02:15
Subject Sample	# 1	0.175	02:16
Breath Volume		3.894	02:16
Air Blank		0.000	02:16
Air Blank		0.000	02:18
Subject Sample	# 2	0.171	02:18
Breath Volume		3.406	02:18
Air Blank		0.000	02:19
Control Test		0.077	02:20
Air Blank		0.000	02:20
Diagnostics Check		OK	02:20

Data Download Date/Time: 04/22/2016 12:13

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORES,ALAN,P
BREATH TEST OPERATOR AGENCY: Holmes Beach PD
ARRESTING OFFICER: BORES,ALAN
ARRESTING OFFICER AGENCY: Holmes Beach PD
LAST AGENCY INSPECTION: 03/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 20:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: PUPEK

DATE/TIME OF TEST: 04/01/2016 23:54:03

Diagnostics Check		OK	23:51	
Air Blank		0.000	23:51	
Control Test		0.079	23:51	
Air Blank		0.000	23:52	
Subject Sample	# 1		23:52	Subject Test Refused
Breath Volume		0.000	23:52	
Air Blank		0.000	23:53	
Control Test		0.079	23:53	
Air Blank		0.000	23:53	
Diagnostics Check		OK	23:54	

Data Download Date/Time: 04/22/2016 12:13

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TATE,ROBERT,J
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: TATE,ROBERT
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 03/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 01:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: MONSALVE

DATE/TIME OF TEST: 04/02/2016 02:09:43

Diagnostics Check		OK	02:00	
Air Blank		0.000	02:00	
Control Test		0.076	02:01	
Air Blank		0.000	02:01	
Subject Sample	# 1	0.252	02:02	
Breath Volume		1.906	02:02	
Air Blank		0.000	02:03	
Air Blank		0.000	02:05	
Subject Sample	# 2	0.251	02:07	
Breath Volume		2.523	02:07	
Air Blank		0.000	02:08	
Control Test		0.073	02:09	Control Outside Tolerance
Air Blank		0.000	02:09	

Data Download Date/Time: 04/22/2016 12:13

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TATE,ROBERT,J
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: TATE,ROBERT
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 03/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 01:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: MONSALVE

DATE/TIME OF TEST: 04/02/2016 02:17:50

Diagnostics Check		OK	02:11
Air Blank		0.000	02:11
Control Test		0.075	02:11
Air Blank		0.000	02:12
Subject Sample	# 1	0.269	02:13
Breath Volume		1.933	02:13
Air Blank		0.000	02:13
Air Blank		0.000	02:15
Subject Sample	# 2	0.250	02:16
Breath Volume		1.496	02:16
Air Blank		0.000	02:16
Control Test		0.076	02:17
Air Blank		0.000	02:17
Diagnostics Check		OK	02:17

Data Download Date/Time: 04/22/2016 12:13

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARTIGAS,KRISTINA,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: ARTIGAS,KRISTINA
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 03/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 00:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:46

SUBJECT'S LAST NAME: BROUSSEAU

DATE/TIME OF TEST: 04/02/2016 02:27:50

Diagnostics Check		OK	02:20
Air Blank		0.000	02:21
Control Test		0.076	02:21
Air Blank		0.000	02:21
Subject Sample	# 1	0.154	02:22
Breath Volume		3.152	02:22
Air Blank		0.000	02:23
Air Blank		0.000	02:25
Subject Sample	# 2	0.150	02:26
Breath Volume		3.550	02:26
Air Blank		0.000	02:26
Control Test		0.077	02:27
Air Blank		0.000	02:27
Diagnostics Check		OK	02:27

Data Download Date/Time: 04/22/2016 12:13

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALLACE,JENNIFER,L
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: WALLACE,JENNIFER
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 03/28/2016

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 03:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:13

SUBJECT'S LAST NAME: DAILY

DATE/TIME OF TEST: 04/03/2016 04:39:55

Diagnostics Check		OK	04:34	
Air Blank		0.000	04:34	
Control Test		0.079	04:35	
Air Blank		0.000	04:35	
Subject Sample	# 1	0.000	04:36	
Breath Volume		2.785	04:36	
Air Blank		0.000	04:36	
Air Blank			04:38	Improper Sample
Air Blank		0.000	04:38	
Control Test		0.079	04:39	
Air Blank		0.000	04:39	
Diagnostics Check		OK	04:39	

Data Download Date/Time: 04/22/2016 12:13

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALLACE,JENNIFER,L
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: WALLACE,JENNIFER
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 03/28/2016

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 03:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:13

SUBJECT'S LAST NAME: DAILY

DATE/TIME OF TEST: 04/03/2016 04:35:22

Diagnostics Check	OK	04:34	
Air Blank		04:34	RFI Detect
Air Blank	0.000	04:35	

Data Download Date/Time: 04/22/2016 12:13

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALLACE,JENNIFER,L
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: WALLACE,JENNIFER
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 03/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 03:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:13

SUBJECT'S LAST NAME: DAILY

DATE/TIME OF TEST: 04/03/2016 04:43:41

Diagnostics Check		OK	04:37
Air Blank		0.000	04:37
Control Test		0.079	04:38
Air Blank		0.000	04:38
Subject Sample	# 1	0.000	04:39
Breath Volume		2.996	04:39
Air Blank		0.000	04:39
Air Blank		0.000	04:41
Subject Sample	# 2	0.000	04:42
Breath Volume		3.253	04:42
Air Blank		0.000	04:42
Control Test		0.079	04:43
Air Blank		0.000	04:43
Diagnostics Check		OK	04:43

Data Download Date/Time: 04/22/2016 12:13

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TATE,ROBERT,J
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: TATE,ROBERT
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 03/28/2016

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 16:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:50

SUBJECT'S LAST NAME: ALVARES

DATE/TIME OF TEST: 04/03/2016 18:33:23

Diagnostics Check		OK	18:26
Air Blank		0.000	18:27
Control Test		0.078	18:27
Air Blank		0.000	18:28
Subject Sample	# 1	0.017	18:28
Breath Volume		3.515	18:28
Air Blank		0.000	18:29
Air Blank		0.000	18:31
Subject Sample	# 2	0.016	18:31
Breath Volume		4.074	18:31
Air Blank		0.000	18:32
Control Test		0.079	18:32
Air Blank		0.000	18:33
Diagnostics Check		OK	18:33

Data Download Date/Time: 04/22/2016 12:13

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAKAI,MICHAEL,T
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: MAKAI,MICHAEL
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 03/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 00:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:28

SUBJECT'S LAST NAME: DICKERSON

DATE/TIME OF TEST: 04/05/2016 01:56:23

Diagnostics Check		OK	01:50
Air Blank		0.000	01:50
Control Test		0.079	01:50
Air Blank		0.000	01:51
Subject Sample	# 1	0.039	01:51
Breath Volume		2.835	01:51
Air Blank		0.000	01:52
Air Blank		0.000	01:54
Subject Sample	# 2	0.038	01:54
Breath Volume		3.097	01:54
Air Blank		0.000	01:55
Control Test		0.079	01:55
Air Blank		0.000	01:56
Diagnostics Check		OK	01:56

Data Download Date/Time: 04/22/2016 12:13

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TATE,ROBERT,J
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: TATE,ROBERT
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 03/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 22:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:27

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 04/07/2016 00:02:13

Diagnostics Check		OK	23:47	
Air Blank		0.000	23:47	
Control Test		0.077	23:48	
Air Blank		0.000	23:48	
Subject Sample	# 2	0.183	23:49	
Breath Volume		2.664	23:49	
Air Blank		0.000	23:50	
Air Blank		0.000	23:52	
Subject Sample	# 3		23:55	No Sample Provided
Breath Volume		0.000	23:55	
Air Blank		0.000	23:55	
Air Blank		0.000	23:57	
Subject Sample	# 1		00:00	No Sample Provided
Breath Volume		0.000	00:00	
Air Blank		0.000	00:01	
Control Test		0.078	00:01	
Air Blank		0.000	00:02	
Diagnostics Check		OK	00:02	

Data Download Date/Time: 04/22/2016 12:13

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TATE,ROBERT,J
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: TATE,ROBERT
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 03/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 22:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:14

SUBJECT'S LAST NAME: BUTLER

DATE/TIME OF TEST: 04/07/2016 23:41:37

Diagnostics Check		OK	23:35
Air Blank		0.000	23:35
Control Test		0.076	23:35
Air Blank		0.000	23:36
Subject Sample	# 1	0.144	23:36
Breath Volume		2.617	23:36
Air Blank		0.000	23:37
Air Blank		0.000	23:39
Subject Sample	# 2	0.147	23:39
Breath Volume		1.363	23:39
Air Blank		0.000	23:40
Control Test		0.077	23:41
Air Blank		0.000	23:41
Diagnostics Check		OK	23:41

Data Download Date/Time: 04/22/2016 12:13

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 03/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 01:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: QUILES FIGUEROA

DATE/TIME OF TEST: 04/09/2016 02:42:09

Diagnostics Check		OK	02:35
Air Blank		0.000	02:35
Control Test		0.078	02:36
Air Blank		0.000	02:36
Subject Sample	# 1	0.127	02:37
Breath Volume		2.371	02:37
Air Blank		0.000	02:38
Air Blank		0.000	02:39
Subject Sample	# 2	0.125	02:40
Breath Volume		3.156	02:40
Air Blank		0.000	02:41
Control Test		0.076	02:41
Air Blank		0.000	02:42
Diagnostics Check		OK	02:42

Data Download Date/Time: 04/22/2016 12:13

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 03/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 23:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: DULIN

DATE/TIME OF TEST: 04/10/2016 00:47:11

Diagnostics Check		OK	00:40
Air Blank		0.000	00:41
Control Test		0.080	00:41
Air Blank		0.000	00:42
Subject Sample	# 1	0.146	00:42
Breath Volume		2.597	00:42
Air Blank		0.000	00:43
Air Blank		0.000	00:45
Subject Sample	# 2	0.152	00:45
Breath Volume		3.359	00:45
Air Blank		0.000	00:46
Control Test		0.078	00:46
Air Blank		0.000	00:47
Diagnostics Check		OK	00:47

Data Download Date/Time: 04/22/2016 12:13

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCGREDE,JOHN,B
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: MCGREDE,JOHN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 03/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 00:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: OSBORNE

DATE/TIME OF TEST: 04/10/2016 01:19:05

Diagnostics Check		OK	01:11
Air Blank		0.000	01:12
Control Test		0.079	01:12
Air Blank		0.000	01:12
Subject Sample	# 1	0.133	01:14
Breath Volume		1.304	01:14
Air Blank		0.000	01:15
Air Blank		0.000	01:16
Subject Sample	# 2	0.128	01:17
Breath Volume		1.472	01:17
Air Blank		0.000	01:18
Control Test		0.076	01:18
Air Blank		0.000	01:19
Diagnostics Check		OK	01:19

Data Download Date/Time: 04/22/2016 12:13

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAKAI,MICHAEL,T
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: MAKAI,MICHAEL
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 03/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 00:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:28

SUBJECT'S LAST NAME: DAY

DATE/TIME OF TEST: 04/10/2016 01:56:17

Diagnostics Check		OK	01:49
Air Blank		0.000	01:49
Control Test		0.078	01:49
Air Blank		0.000	01:50
Subject Sample	# 1	0.000	01:50
Breath Volume		1.882	01:50
Air Blank		0.000	01:51
Air Blank		0.000	01:53
Subject Sample	# 2	0.000	01:54
Breath Volume		2.074	01:54
Air Blank		0.000	01:55
Control Test		0.080	01:55
Air Blank		0.000	01:56
Diagnostics Check		OK	01:56

Data Download Date/Time: 04/22/2016 12:13

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 03/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 04:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: SCOTT

DATE/TIME OF TEST: 04/10/2016 04:57:46

Diagnostics Check		OK	04:50
Air Blank		0.000	04:51
Control Test		0.079	04:51
Air Blank		0.000	04:52
Subject Sample	# 1	0.176	04:52
Breath Volume		1.472	04:52
Air Blank		0.000	04:53
Air Blank		0.000	04:55
Subject Sample	# 2	0.167	04:56
Breath Volume		1.734	04:56
Air Blank		0.000	04:56
Control Test		0.077	04:57
Air Blank		0.000	04:57
Diagnostics Check		OK	04:57

Data Download Date/Time: 04/22/2016 12:13

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 03/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 21:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:18

SUBJECT'S LAST NAME: MARTIN

DATE/TIME OF TEST: 04/10/2016 22:44:51

Diagnostics Check		OK	22:38
Air Blank		0.000	22:39
Control Test		0.080	22:39
Air Blank		0.000	22:39
Subject Sample	# 1	0.200	22:40
Breath Volume		2.093	22:40
Air Blank		0.000	22:41
Air Blank		0.000	22:42
Subject Sample	# 2	0.198	22:43
Breath Volume		2.304	22:43
Air Blank		0.000	22:43
Control Test		0.077	22:44
Air Blank		0.000	22:44
Diagnostics Check		OK	22:44

Data Download Date/Time: 04/22/2016 12:13

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORES,ALAN,P
BREATH TEST OPERATOR AGENCY: Holmes Beach PD
ARRESTING OFFICER: BORES,ALAN
ARRESTING OFFICER AGENCY: Holmes Beach PD
LAST AGENCY INSPECTION: 03/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 03:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:48

SUBJECT'S LAST NAME: CANNON

DATE/TIME OF TEST: 04/12/2016 05:16:28

Diagnostics Check		OK	05:09
Air Blank		0.000	05:10
Control Test		0.079	05:10
Air Blank		0.000	05:11
Subject Sample	# 1	0.108	05:11
Breath Volume		1.277	05:11
Air Blank		0.000	05:12
Air Blank		0.000	05:14
Subject Sample	# 2	0.111	05:14
Breath Volume		1.656	05:14
Air Blank		0.000	05:15
Control Test		0.079	05:15
Air Blank		0.000	05:16
Diagnostics Check		OK	05:16

Data Download Date/Time: 04/22/2016 12:13

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLINE,GLENN,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: CLINE,GLENN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 03/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: SYS CHECK
ARREST TIME: 13:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:00

SUBJECT'S LAST NAME: TEST

DATE/TIME OF TEST: 04/20/2016 15:05:29

Diagnostics Check		OK	14:59
Air Blank		0.000	15:00
Control Test		0.079	15:00
Air Blank		0.000	15:01
Subject Sample	# 1	0.000	15:01
Breath Volume		1.785	15:01
Air Blank		0.000	15:01
Air Blank		0.000	15:03
Subject Sample	# 2	0.000	15:04
Breath Volume		1.585	15:04
Air Blank		0.000	15:04
Control Test		0.079	15:04
Air Blank		0.000	15:05
Diagnostics Check		OK	15:05

Data Download Date/Time: 04/22/2016 12:13

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 03/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 00:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:09

SUBJECT'S LAST NAME: HENDRIX

DATE/TIME OF TEST: 04/14/2016 01:36:01

Diagnostics Check		OK	01:29
Air Blank		0.000	01:30
Control Test		0.077	01:30
Air Blank		0.000	01:31
Subject Sample	# 1	0.168	01:31
Breath Volume		3.945	01:31
Air Blank		0.000	01:32
Air Blank		0.000	01:33
Subject Sample	# 2	0.171	01:34
Breath Volume		3.257	01:34
Air Blank		0.000	01:35
Control Test		0.076	01:35
Air Blank		0.000	01:35
Diagnostics Check		OK	01:36

Data Download Date/Time: 04/22/2016 12:39

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDERMOTT,JAMES,W
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: MCDERMOTT,JAMES
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 03/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 00:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: DARNELL

DATE/TIME OF TEST: 04/15/2016 02:21:50

Diagnostics Check		OK	02:18	
Air Blank		0.000	02:19	
Control Test		0.078	02:19	
Air Blank		0.000	02:20	
Subject Sample	# 1		02:20	Subject Test Refused
Breath Volume		0.000	02:20	
Air Blank		0.000	02:20	
Control Test		0.078	02:21	
Air Blank		0.000	02:21	
Diagnostics Check		OK	02:21	

Data Download Date/Time: 04/22/2016 12:39

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCGREDE,JOHN,B
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: MCGREDE,JOHN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 03/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 03:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: ESPINOZA LARA

DATE/TIME OF TEST: 04/16/2016 03:57:28

Diagnostics Check		OK	03:50
Air Blank		0.000	03:51
Control Test		0.077	03:51
Air Blank		0.000	03:51
Subject Sample	# 1	0.090	03:52
Breath Volume		2.695	03:52
Air Blank		0.000	03:53
Air Blank		0.000	03:55
Subject Sample	# 2	0.091	03:55
Breath Volume		4.859	03:55
Air Blank		0.000	03:56
Control Test		0.076	03:56
Air Blank		0.000	03:57
Diagnostics Check		OK	03:57

Data Download Date/Time: 04/22/2016 12:39

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KENYAN,M,A
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: KENYAN,M
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 03/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 03:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:50

SUBJECT'S LAST NAME: AUSTIN

DATE/TIME OF TEST: 04/16/2016 05:28:09

Diagnostics Check		OK	05:14	
Air Blank		0.000	05:15	
Control Test		0.078	05:15	
Air Blank		0.000	05:16	
Subject Sample	# 1	0.105	05:19	Volume Not Met
Breath Volume		0.292	05:19	
Air Blank		0.000	05:20	
Air Blank		0.000	05:21	
Subject Sample	# 2	0.142	05:23	
Breath Volume		2.085	05:23	
Air Blank		0.000	05:24	
Air Blank		0.000	05:26	
Subject Sample	# 3	0.136	05:26	
Breath Volume		2.656	05:26	
Air Blank		0.000	05:27	
Control Test		0.076	05:27	
Air Blank		0.000	05:28	
Diagnostics Check		OK	05:28	

Data Download Date/Time: 04/22/2016 12:39

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TATE,ROBERT,J
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: TATE,ROBERT
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 03/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 17:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:58

SUBJECT'S LAST NAME: O'HALLORAN

DATE/TIME OF TEST: 04/16/2016 18:26:26

Diagnostics Check		OK	18:19
Air Blank		0.000	18:19
Control Test		0.076	18:19
Air Blank		0.000	18:20
Subject Sample	# 1	0.194	18:21
Breath Volume		1.937	18:21
Air Blank		0.000	18:22
Air Blank		0.000	18:24
Subject Sample	# 2	0.193	18:24
Breath Volume		1.699	18:24
Air Blank		0.000	18:25
Control Test		0.075	18:25
Air Blank		0.000	18:26
Diagnostics Check		OK	18:26

Data Download Date/Time: 04/22/2016 12:40

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TATE,ROBERT,J
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: TATE,ROBERT
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 03/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 01:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: OKOLOWITCZ

DATE/TIME OF TEST: 04/17/2016 02:18:40

Diagnostics Check		OK	02:10
Air Blank		0.000	02:10
Control Test		0.077	02:11
Air Blank		0.000	02:11
Subject Sample	# 1	0.152	02:13
Breath Volume		2.214	02:13
Air Blank		0.000	02:14
Air Blank		0.000	02:15
Subject Sample	# 2	0.142	02:17
Breath Volume		1.515	02:17
Air Blank		0.000	02:17
Control Test		0.076	02:18
Air Blank		0.000	02:18
Diagnostics Check		OK	02:18

Data Download Date/Time: 04/22/2016 12:40

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCGREDE,JOHN,B
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: MCGREDE,JOHN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 03/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 01:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: MARCIANO

DATE/TIME OF TEST: 04/17/2016 02:59:35

Diagnostics Check		OK	02:56	
Air Blank		0.000	02:56	
Control Test		0.078	02:57	
Air Blank		0.000	02:57	
Subject Sample	# 1		02:58	Subject Test Refused
Breath Volume		0.000	02:58	
Air Blank		0.000	02:58	
Control Test		0.078	02:58	
Air Blank		0.000	02:59	
Diagnostics Check		OK	02:59	

Data Download Date/Time: 04/22/2016 12:40

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORES,ALAN,P
BREATH TEST OPERATOR AGENCY: Holmes Beach PD
ARRESTING OFFICER: BORES,ALAN
ARRESTING OFFICER AGENCY: Holmes Beach PD
LAST AGENCY INSPECTION: 03/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 04:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:20

SUBJECT'S LAST NAME: RICHMAN

DATE/TIME OF TEST: 04/17/2016 06:51:49

Diagnostics Check		OK	06:43
Air Blank		0.000	06:44
Control Test		0.079	06:44
Air Blank		0.000	06:45
Subject Sample	# 1	0.030	06:46
Breath Volume		1.394	06:46
Air Blank		0.000	06:47
Air Blank		0.000	06:48
Subject Sample	# 2	0.028	06:50
Breath Volume		1.308	06:50
Air Blank		0.000	06:50
Control Test		0.079	06:51
Air Blank		0.000	06:51
Diagnostics Check		OK	06:51

Data Download Date/Time: 04/22/2016 12:40

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORES,ALAN,P
BREATH TEST OPERATOR AGENCY: Holmes Beach PD
ARRESTING OFFICER: BORES,ALAN
ARRESTING OFFICER AGENCY: Holmes Beach PD
LAST AGENCY INSPECTION: 03/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:38

SUBJECT'S LAST NAME: CUELLAR

DATE/TIME OF TEST: 04/18/2016 03:13:59

Diagnostics Check		OK	03:00	
Air Blank		0.000	03:01	
Control Test		0.079	03:01	
Air Blank		0.000	03:02	
Subject Sample	# 1	0.137	03:03	
Breath Volume		3.039	03:03	
Air Blank		0.000	03:04	
Air Blank		0.000	03:05	
Subject Sample	# 2		03:08	No Sample Provided
Breath Volume		0.000	03:08	
Air Blank		0.000	03:09	
Air Blank		0.000	03:11	
Subject Sample	# 3	0.131	03:12	
Breath Volume		3.792	03:12	
Air Blank		0.000	03:13	
Control Test		0.079	03:13	
Air Blank		0.000	03:13	
Diagnostics Check		OK	03:13	

Data Download Date/Time: 04/22/2016 12:40

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 03/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 22:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: SYRING

DATE/TIME OF TEST: 04/19/2016 23:29:16

Diagnostics Check		OK	23:26	
Air Blank		0.000	23:26	
Control Test		0.078	23:26	
Air Blank		0.000	23:27	
Subject Sample	# 1		23:27	Subject Test Refused
Breath Volume		0.000	23:27	
Air Blank		0.000	23:28	
Control Test		0.078	23:28	
Air Blank		0.000	23:29	
Diagnostics Check		OK	23:29	

Data Download Date/Time: 04/22/2016 12:40

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLINE,GLENN,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: CLINE,GLENN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 03/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: SYS CHECK
ARREST TIME: 13:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:00

SUBJECT'S LAST NAME: TEST

DATE/TIME OF TEST: 04/20/2016 14:49:24

Diagnostics Check		OK	14:43
Air Blank		0.000	14:43
Control Test		0.078	14:44
Air Blank		0.000	14:44
Subject Sample	# 1	0.000	14:45
Breath Volume		1.984	14:45
Air Blank		0.000	14:45
Air Blank		0.000	14:47
Subject Sample	# 2	0.000	14:48
Breath Volume		1.695	14:48
Air Blank		0.000	14:48
Control Test		0.079	14:48
Air Blank		0.000	14:49
Diagnostics Check		OK	14:49

Data Download Date/Time: 04/22/2016 12:40

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Headquarters
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Headquarters
LAST AGENCY INSPECTION: 04/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 10:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:24

SUBJECT'S LAST NAME: GARCIA MARTINEZ

DATE/TIME OF TEST: 04/22/2016 23:50:40

Diagnostics Check		OK	23:44
Air Blank		0.000	23:44
Control Test		0.077	23:45
Air Blank		0.000	23:45
Subject Sample	# 1	0.127	23:46
Breath Volume		2.164	23:46
Air Blank		0.000	23:46
Air Blank		0.000	23:48
Subject Sample	# 2	0.125	23:48
Breath Volume		2.601	23:48
Air Blank		0.000	23:49
Control Test		0.076	23:50
Air Blank		0.000	23:50
Diagnostics Check		OK	23:50

Data Download Date/Time: 05/25/2016 09:29

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 04/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 02:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:54

SUBJECT'S LAST NAME: ALBILLANEZ-RUIZ

DATE/TIME OF TEST: 04/23/2016 03:20:51

Diagnostics Check		OK	03:14
Air Blank		0.000	03:14
Control Test		0.078	03:15
Air Blank		0.000	03:15
Subject Sample	# 1	0.196	03:16
Breath Volume		1.890	03:16
Air Blank		0.000	03:17
Air Blank		0.000	03:18
Subject Sample	# 2	0.190	03:19
Breath Volume		1.960	03:19
Air Blank		0.000	03:19
Control Test		0.077	03:20
Air Blank		0.000	03:20
Diagnostics Check		OK	03:20

Data Download Date/Time: 05/25/2016 09:29

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAKAI,MICHAEL,T
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: MAKAI,MICHAEL
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 04/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 03:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:42

SUBJECT'S LAST NAME: HUGHES

DATE/TIME OF TEST: 04/23/2016 04:13:12

Diagnostics Check		OK	04:10	
Air Blank		0.000	04:10	
Control Test		0.078	04:11	
Air Blank		0.000	04:11	
Subject Sample	# 1		04:11	Subject Test Refused
Breath Volume		0.000	04:11	
Air Blank		0.000	04:12	
Control Test		0.078	04:12	
Air Blank		0.000	04:13	
Diagnostics Check		OK	04:13	

Data Download Date/Time: 05/25/2016 09:29

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAKAI,MICHAEL,T
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: MAKAI,MICHAEL
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 04/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 21:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:58

SUBJECT'S LAST NAME: HOUSE

DATE/TIME OF TEST: 04/23/2016 22:21:42

Diagnostics Check		OK	22:18	
Air Blank		0.000	22:19	
Control Test		0.077	22:19	
Air Blank		0.000	22:19	
Subject Sample	# 1		22:20	Subject Test Refused
Breath Volume		0.000	22:20	
Air Blank		0.000	22:20	
Control Test		0.078	22:21	
Air Blank		0.000	22:21	
Diagnostics Check		OK	22:21	

Data Download Date/Time: 05/25/2016 09:29

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 04/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 21:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 04/23/2016 23:12:14

Diagnostics Check		OK	23:04
Air Blank		0.000	23:04
Control Test		0.077	23:05
Air Blank		0.000	23:05
Subject Sample	# 1	0.134	23:06
Breath Volume		1.207	23:06
Air Blank		0.000	23:07
Air Blank		0.000	23:09
Subject Sample	# 2	0.129	23:10
Breath Volume		1.644	23:10
Air Blank		0.000	23:11
Control Test		0.075	23:11
Air Blank		0.000	23:12
Diagnostics Check		OK	23:12

Data Download Date/Time: 05/25/2016 09:29

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 04/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 11:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:48

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 04/24/2016 01:17:40

Diagnostics Check		OK	01:08
Air Blank		0.000	01:09
Control Test		0.078	01:09
Air Blank		0.000	01:10
Subject Sample	# 1	0.117	01:11
Breath Volume		1.515	01:11
Air Blank		0.000	01:12
Air Blank		0.000	01:14
Subject Sample	# 2	0.119	01:15
Breath Volume		2.359	01:15
Air Blank		0.000	01:16
Control Test		0.076	01:17
Air Blank		0.000	01:17
Diagnostics Check		OK	01:17

Data Download Date/Time: 05/25/2016 09:29

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 04/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 00:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:33

SUBJECT'S LAST NAME: PEELE

DATE/TIME OF TEST: 04/24/2016 02:00:59

Diagnostics Check		OK	01:53	
Air Blank		0.000	01:54	
Control Test		0.078	01:54	
Air Blank		0.000	01:55	
Subject Sample	# 1	0.143	01:55	
Breath Volume		2.800	01:55	
Air Blank		0.000	01:56	
Air Blank		0.000	01:58	
Subject Sample	# 2	0.128	01:59	
Breath Volume		1.582	01:59	
Air Blank			02:00	Purge Fail
Air Blank		0.000	02:00	

Data Download Date/Time: 05/25/2016 09:29

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 04/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 00:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:33

SUBJECT'S LAST NAME: PEELE

DATE/TIME OF TEST: 04/24/2016 02:10:37

Diagnostics Check		OK	02:04
Air Blank		0.000	02:04
Control Test		0.075	02:04
Air Blank		0.000	02:05
Subject Sample	# 1	0.145	02:06
Breath Volume		2.476	02:06
Air Blank		0.000	02:06
Air Blank		0.000	02:08
Subject Sample	# 2	0.125	02:08
Breath Volume		1.703	02:08
Air Blank		0.000	02:09
Control Test		0.077	02:10
Air Blank		0.000	02:10
Diagnostics Check		OK	02:10

Data Download Date/Time: 05/25/2016 09:29

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAKAI,MICHAEL,T
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: MAKAI,MICHAEL
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 04/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 01:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:04

SUBJECT'S LAST NAME: BISCHE

DATE/TIME OF TEST: 04/24/2016 02:34:29

Diagnostics Check		OK	02:24
Air Blank		0.000	02:25
Control Test		0.077	02:25
Air Blank		0.000	02:26
Subject Sample	# 1	0.134	02:26
Breath Volume		1.312	02:26
Air Blank		0.000	02:27
Air Blank		0.000	02:29
Subject Sample	# 2	0.158	02:29
Breath Volume		1.671	02:29
Air Blank		0.000	02:30
Air Blank		0.000	02:32
Subject Sample	# 3	0.154	02:32
Breath Volume		2.445	02:32
Air Blank		0.000	02:33
Control Test		0.077	02:33
Air Blank		0.000	02:34
Diagnostics Check		OK	02:34

Data Download Date/Time: 05/25/2016 09:29

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 04/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 00:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:44

SUBJECT'S LAST NAME: MENDEZ

DATE/TIME OF TEST: 04/25/2016 01:25:47

Diagnostics Check		OK	01:13	
Air Blank		0.000	01:14	
Control Test		0.078	01:14	
Air Blank		0.000	01:15	
Subject Sample	# 1	0.147	01:18	Volume Not Met
Breath Volume		0.843	01:18	
Air Blank		0.000	01:19	
Air Blank		0.000	01:20	
Subject Sample	# 2	0.136	01:24	Volume Not Met
Breath Volume		0.281	01:24	
Air Blank		0.000	01:24	
Control Test		0.076	01:25	
Air Blank		0.000	01:25	
Diagnostics Check		OK	01:25	

Data Download Date/Time: 05/25/2016 09:29

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 04/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 00:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:44

SUBJECT'S LAST NAME: MENDEZ

DATE/TIME OF TEST: 04/25/2016 01:41:12

Diagnostics Check		OK	01:27	
Air Blank		0.000	01:27	
Control Test		0.076	01:27	
Air Blank		0.000	01:28	
Subject Sample	# 1	0.119	01:31	Volume Not Met
Breath Volume		0.582	01:31	
Air Blank		0.000	01:32	
Air Blank		0.000	01:34	
Subject Sample	# 2	0.161	01:34	
Breath Volume		1.722	01:34	
Air Blank		0.000	01:35	
Air Blank		0.000	01:37	
Subject Sample	# 3	0.150	01:39	
Breath Volume		1.722	01:39	
Air Blank		0.000	01:40	
Control Test		0.074	01:40	Control Outside Tolerance
Air Blank		0.000	01:41	

Data Download Date/Time: 05/25/2016 09:29

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VONADOR,JERARD,E
BREATH TEST OPERATOR AGENCY: Bradenton PD
ARRESTING OFFICER: VONADOR,JERARD
ARRESTING OFFICER AGENCY: Bradenton PD
LAST AGENCY INSPECTION: 04/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 00:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:46

SUBJECT'S LAST NAME: RACHIDE

DATE/TIME OF TEST: 04/25/2016 01:52:03

Diagnostics Check		OK	01:45
Air Blank		0.000	01:45
Control Test		0.076	01:45
Air Blank		0.000	01:46
Subject Sample	# 1	0.304	01:47
Breath Volume		2.234	01:47
Air Blank		0.000	01:48
Air Blank		0.000	01:49
Subject Sample	# 2	0.304	01:50
Breath Volume		2.949	01:50
Air Blank		0.000	01:51
Control Test		0.076	01:51
Air Blank		0.000	01:51
Diagnostics Check		OK	01:52

Data Download Date/Time: 05/25/2016 09:29

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 04/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 00:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:44

SUBJECT'S LAST NAME: MENDEZ

DATE/TIME OF TEST: 04/25/2016 02:03:29

Diagnostics Check		OK	01:54	
Air Blank		0.000	01:54	
Control Test		0.077	01:55	
Air Blank		0.000	01:55	
Subject Sample	# 1	0.145	01:58	Volume Not Met
Breath Volume		0.128	01:58	
Air Blank		0.000	01:59	
Air Blank		0.000	02:01	
Subject Sample	# 2		02:02	Subject Test Refused
Breath Volume		0.000	02:02	
Air Blank		0.000	02:02	
Control Test		0.073	02:02	Control Outside Tolerance
Air Blank		0.000	02:03	

Data Download Date/Time: 05/25/2016 09:29

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 04/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 05:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:52

SUBJECT'S LAST NAME: GANPAT

DATE/TIME OF TEST: 04/25/2016 06:19:36

Diagnostics Check		OK	06:12
Air Blank		0.000	06:13
Control Test		0.077	06:13
Air Blank		0.000	06:14
Subject Sample	# 1	0.128	06:14
Breath Volume		1.984	06:14
Air Blank		0.000	06:15
Air Blank		0.000	06:17
Subject Sample	# 2	0.134	06:17
Breath Volume		1.289	06:17
Air Blank		0.000	06:18
Control Test		0.076	06:18
Air Blank		0.000	06:19
Diagnostics Check		OK	06:19

Data Download Date/Time: 05/25/2016 09:30

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAKAI,MICHAEL,T
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: MAKAI,MICHAEL
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 04/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 01:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: WESTBROOK

DATE/TIME OF TEST: 04/28/2016 03:22:10

Diagnostics Check		OK	03:15
Air Blank		0.000	03:15
Control Test		0.079	03:15
Air Blank		0.000	03:16
Subject Sample	# 1	0.251	03:16
Breath Volume		2.527	03:16
Air Blank		0.000	03:17
Air Blank		0.000	03:19
Subject Sample	# 2	0.247	03:20
Breath Volume		2.710	03:20
Air Blank		0.000	03:21
Control Test		0.078	03:21
Air Blank		0.000	03:22
Diagnostics Check		OK	03:22

Data Download Date/Time: 05/25/2016 09:30

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAKAI,MICHAEL,T
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: MAKAI,MICHAEL
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 04/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 22:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: GLIDDEN

DATE/TIME OF TEST: 04/28/2016 23:15:54

Diagnostics Check		OK	23:06	
Air Blank		0.000	23:06	
Control Test		0.078	23:07	
Air Blank		0.000	23:07	
Subject Sample	# 1		23:10	No Sample Provided
Breath Volume		0.000	23:10	
Air Blank		0.000	23:11	
Air Blank		0.000	23:13	
Subject Sample	# 2		23:14	Subject Test Refused
Breath Volume		0.000	23:14	
Air Blank		0.000	23:14	
Control Test		0.078	23:15	
Air Blank		0.000	23:15	
Diagnostics Check		OK	23:15	

Data Download Date/Time: 05/25/2016 09:30

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TATE,ROBERT,J
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: TATE,ROBERT
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 04/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 19:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:35

SUBJECT'S LAST NAME: BEDWELL

DATE/TIME OF TEST: 04/30/2016 21:08:09

Diagnostics Check		OK	21:04	
Air Blank		0.000	21:05	
Control Test		0.078	21:05	
Air Blank		0.000	21:06	
Subject Sample	# 1		21:06	Subject Test Refused
Breath Volume		0.000	21:06	
Air Blank		0.000	21:07	
Control Test		0.077	21:07	
Air Blank		0.000	21:08	
Diagnostics Check		OK	21:08	

Data Download Date/Time: 05/25/2016 09:30

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK,CASEY,S
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VAN-TREESE,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 03/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A1
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:39

SUBJECT'S LAST NAME: VIANA

DATE/TIME OF TEST: 04/01/2016 03:06:58

Diagnostics Check		OK	03:00
Air Blank		0.000	03:00
Control Test		0.078	03:00
Air Blank		0.000	03:01
Subject Sample	# 1	0.200	03:02
Breath Volume		2.007	03:02
Air Blank		0.000	03:03
Air Blank		0.000	03:04
Subject Sample	# 2	0.199	03:05
Breath Volume		1.210	03:05
Air Blank		0.000	03:06
Control Test		0.077	03:06
Air Blank		0.000	03:06
Diagnostics Check		OK	03:06

Data Download Date/Time: 04/06/2016 19:42

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BADEN,T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 03/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A1
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: GALAL

DATE/TIME OF TEST: 04/01/2016 04:26:44

Diagnostics Check		OK	04:20
Air Blank		0.000	04:20
Control Test		0.078	04:21
Air Blank		0.000	04:21
Subject Sample	# 1	0.149	04:22
Breath Volume		1.746	04:22
Air Blank		0.000	04:22
Air Blank		0.000	04:24
Subject Sample	# 2	0.163	04:25
Breath Volume		1.281	04:25
Air Blank		0.000	04:25
Control Test		0.078	04:26
Air Blank		0.000	04:26
Diagnostics Check		OK	04:26

Data Download Date/Time: 04/06/2016 19:43

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKIPPER,ROGER,G
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARBOSA,F 924/27097
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 03/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A1
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 07:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:28

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 04/01/2016 09:01:54

Diagnostics Check		OK	08:54
Air Blank		0.000	08:55
Control Test		0.078	08:55
Air Blank		0.000	08:56
Subject Sample	# 1	0.128	08:57
Breath Volume		1.878	08:57
Air Blank		0.000	08:57
Air Blank		0.000	08:59
Subject Sample	# 2	0.130	09:00
Breath Volume		1.628	09:00
Air Blank		0.000	09:00
Control Test		0.078	09:01
Air Blank		0.000	09:01
Diagnostics Check		OK	09:01

Data Download Date/Time: 04/06/2016 19:43

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 03/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A1
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:59

SUBJECT'S LAST NAME: FASSO

DATE/TIME OF TEST: 04/01/2016 22:26:10

Diagnostics Check		OK	22:19
Air Blank		0.000	22:20
Control Test		0.079	22:20
Air Blank		0.000	22:21
Subject Sample	# 1	0.095	22:21
Breath Volume		1.214	22:21
Air Blank		0.000	22:22
Air Blank		0.000	22:24
Subject Sample	# 2	0.107	22:24
Breath Volume		2.031	22:24
Air Blank		0.000	22:25
Control Test		0.078	22:25
Air Blank		0.000	22:26
Diagnostics Check		OK	22:26

Data Download Date/Time: 04/06/2016 19:43

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FRANKLIN,R
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 03/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A1
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:19

SUBJECT'S LAST NAME: VELAZQUEZ

DATE/TIME OF TEST: 04/02/2016 01:50:35

Diagnostics Check		OK	01:44
Air Blank		0.000	01:44
Control Test		0.078	01:44
Air Blank		0.000	01:45
Subject Sample	# 1	0.211	01:46
Breath Volume		1.949	01:46
Air Blank		0.000	01:46
Air Blank		0.000	01:48
Subject Sample	# 2	0.214	01:48
Breath Volume		2.230	01:48
Air Blank		0.000	01:49
Control Test		0.077	01:49
Air Blank		0.000	01:50
Diagnostics Check		OK	01:50

Data Download Date/Time: 04/06/2016 19:43

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK,CASEY,S
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 03/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A1
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: QUIROLA

DATE/TIME OF TEST: 04/02/2016 04:36:28

Diagnostics Check		OK	04:30
Air Blank		0.000	04:31
Control Test		0.078	04:31
Air Blank		0.000	04:31
Subject Sample	# 1	0.000	04:32
Breath Volume		1.812	04:32
Air Blank		0.000	04:32
Air Blank		0.000	04:34
Subject Sample	# 2	0.000	04:35
Breath Volume		1.992	04:35
Air Blank		0.000	04:35
Control Test		0.078	04:35
Air Blank		0.000	04:36
Diagnostics Check		OK	04:36

Data Download Date/Time: 04/06/2016 19:43

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SARRO,TAYLOR,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RODRIGUEZ,R
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A1
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:42

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 04/02/2016 23:09:45

Diagnostics Check		OK	23:02
Air Blank		0.000	23:02
Control Test		0.078	23:02
Air Blank		0.000	23:03
Subject Sample	# 1	0.151	23:05
Breath Volume		1.730	23:05
Air Blank		0.000	23:05
Air Blank		0.000	23:07
Subject Sample	# 2	0.149	23:08
Breath Volume		1.765	23:08
Air Blank		0.000	23:08
Control Test		0.077	23:09
Air Blank		0.000	23:09
Diagnostics Check		OK	23:09

Data Download Date/Time: 04/06/2016 19:43

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY,PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 03/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A1
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: SHAH

DATE/TIME OF TEST: 04/03/2016 04:22:44

Diagnostics Check		OK	04:15
Air Blank		0.000	04:15
Control Test		0.078	04:16
Air Blank		0.000	04:16
Subject Sample	# 1	0.188	04:18
Breath Volume		2.582	04:18
Air Blank		0.000	04:18
Air Blank		0.000	04:20
Subject Sample	# 2	0.191	04:21
Breath Volume		2.718	04:21
Air Blank		0.000	04:21
Control Test		0.077	04:22
Air Blank		0.000	04:22
Diagnostics Check		OK	04:22

Data Download Date/Time: 04/06/2016 19:43

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY,PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WINGATE,R
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 03/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A1
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: SCHUCK

DATE/TIME OF TEST: 04/03/2016 04:47:25

Diagnostics Check		OK	04:40
Air Blank		0.000	04:40
Control Test		0.077	04:41
Air Blank		0.000	04:41
Subject Sample	# 1	0.164	04:42
Breath Volume		2.242	04:42
Air Blank		0.000	04:43
Air Blank		0.000	04:45
Subject Sample	# 2	0.167	04:45
Breath Volume		1.578	04:45
Air Blank		0.000	04:46
Control Test		0.077	04:46
Air Blank		0.000	04:47
Diagnostics Check		OK	04:47

Data Download Date/Time: 04/06/2016 19:43

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARLAUG,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 03/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A1
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:35

SUBJECT'S LAST NAME: CASTANON

DATE/TIME OF TEST: 04/03/2016 05:02:34

Diagnostics Check		OK	04:55
Air Blank		0.000	04:56
Control Test		0.078	04:56
Air Blank		0.000	04:57
Subject Sample	# 1	0.155	04:57
Breath Volume		2.902	04:57
Air Blank		0.000	04:58
Air Blank		0.000	05:00
Subject Sample	# 2	0.141	05:00
Breath Volume		1.128	05:00
Air Blank		0.000	05:01
Control Test		0.078	05:01
Air Blank		0.000	05:02
Diagnostics Check		OK	05:02

Data Download Date/Time: 04/06/2016 19:43

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SARRO,TAYLOR,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DEJESUS,V
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 03/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A1
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 05:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:45

SUBJECT'S LAST NAME: ROQUE-RUIZ

DATE/TIME OF TEST: 04/03/2016 06:13:18

Diagnostics Check		OK	06:05
Air Blank		0.000	06:06
Control Test		0.078	06:06
Air Blank		0.000	06:06
Subject Sample	# 1	0.220	06:08
Breath Volume		1.843	06:08
Air Blank		0.000	06:08
Air Blank		0.000	06:10
Subject Sample	# 2	0.219	06:11
Breath Volume		2.277	06:11
Air Blank		0.000	06:12
Control Test		0.077	06:12
Air Blank		0.000	06:13
Diagnostics Check		OK	06:13

Data Download Date/Time: 04/06/2016 19:43

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY,PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARLAUG,G
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 03/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A1
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:29

SUBJECT'S LAST NAME: LOHBERG

DATE/TIME OF TEST: 04/03/2016 23:57:09

Diagnostics Check		OK	23:50
Air Blank		0.000	23:50
Control Test		0.079	23:50
Air Blank		0.000	23:51
Subject Sample	# 1	0.269	23:52
Breath Volume		2.503	23:52
Air Blank		0.000	23:52
Air Blank		0.000	23:54
Subject Sample	# 2	0.268	23:55
Breath Volume		2.347	23:55
Air Blank		0.000	23:56
Control Test		0.077	23:56
Air Blank		0.000	23:57
Diagnostics Check		OK	23:57

Data Download Date/Time: 04/06/2016 19:43

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SARRO,TAYLOR,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARLAUG,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 03/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A1
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:26

SUBJECT'S LAST NAME: HARR

DATE/TIME OF TEST: 04/04/2016 02:54:35

Diagnostics Check		OK	02:46
Air Blank		0.000	02:47
Control Test		0.078	02:47
Air Blank		0.000	02:47
Subject Sample	# 1	0.176	02:49
Breath Volume		2.062	02:49
Air Blank		0.000	02:50
Air Blank		0.000	02:52
Subject Sample	# 2	0.175	02:52
Breath Volume		1.691	02:52
Air Blank		0.000	02:53
Control Test		0.078	02:53
Air Blank		0.000	02:54
Diagnostics Check		OK	02:54

Data Download Date/Time: 04/06/2016 19:43

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MILLER,J
ARRESTING OFFICER AGENCY: Temple Terrace PD
LAST AGENCY INSPECTION: 03/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A1
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:48

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 04/05/2016 01:20:09

Diagnostics Check		OK	01:08	
Air Blank		0.000	01:08	
Control Test		0.079	01:09	
Air Blank		0.000	01:09	
Subject Sample	# 1	0.183	01:12	Volume Not Met
Breath Volume		0.109	01:12	
Air Blank		0.000	01:13	
Air Blank		0.000	01:15	
Subject Sample	# 2	0.215	01:18	Volume Not Met
Breath Volume		0.042	01:18	
Air Blank		0.000	01:19	
Control Test		0.077	01:19	
Air Blank		0.000	01:20	
Diagnostics Check		OK	01:20	

Data Download Date/Time: 04/06/2016 19:43

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MILLER,J
ARRESTING OFFICER AGENCY: Temple Terrace PD
LAST AGENCY INSPECTION: 03/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A1
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:48

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 04/05/2016 01:34:57

Diagnostics Check		OK	01:23	
Air Blank		0.000	01:23	
Control Test		0.077	01:24	
Air Blank		0.000	01:24	
Subject Sample	# 1	0.243	01:27	Volume Not Met
Breath Volume		0.125	01:27	
Air Blank		0.000	01:28	
Air Blank		0.000	01:30	
Subject Sample	# 2		01:33	No Sample Provided
Breath Volume		0.000	01:33	
Air Blank		0.000	01:34	
Control Test		0.076	01:34	
Air Blank		0.000	01:34	
Diagnostics Check		OK	01:34	

Data Download Date/Time: 04/06/2016 19:43

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK,CASEY,S
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WOODEN,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 03/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A1
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: STRONG

DATE/TIME OF TEST: 04/06/2016 00:43:32

Diagnostics Check		OK	00:36
Air Blank		0.000	00:37
Control Test		0.079	00:37
Air Blank		0.000	00:38
Subject Sample	# 1	0.287	00:38
Breath Volume		3.281	00:38
Air Blank		0.000	00:39
Air Blank		0.000	00:41
Subject Sample	# 2	0.281	00:41
Breath Volume		3.277	00:41
Air Blank		0.000	00:42
Control Test		0.077	00:42
Air Blank		0.000	00:43
Diagnostics Check		OK	00:43

Data Download Date/Time: 04/06/2016 19:43

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: AGUAYO,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 03/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A1
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: CALVO-VAZQUEZ

DATE/TIME OF TEST: 04/06/2016 01:31:58

Diagnostics Check		OK	01:25
Air Blank		0.000	01:26
Control Test		0.079	01:26
Air Blank		0.000	01:27
Subject Sample	# 1	0.229	01:27
Breath Volume		3.847	01:27
Air Blank		0.000	01:28
Air Blank		0.000	01:29
Subject Sample	# 2	0.224	01:30
Breath Volume		4.000	01:30
Air Blank		0.000	01:31
Control Test		0.078	01:31
Air Blank		0.000	01:31
Diagnostics Check		OK	01:31

Data Download Date/Time: 04/06/2016 19:43

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK,CASEY,S
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SIMPKINS,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 03/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A1
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:46

SUBJECT'S LAST NAME: HURSH

DATE/TIME OF TEST: 04/06/2016 04:13:16

Diagnostics Check		OK	04:06
Air Blank		0.000	04:07
Control Test		0.078	04:07
Air Blank		0.000	04:08
Subject Sample	# 1	0.147	04:08
Breath Volume		2.984	04:08
Air Blank		0.000	04:09
Air Blank		0.000	04:11
Subject Sample	# 2	0.150	04:11
Breath Volume		3.550	04:11
Air Blank		0.000	04:12
Control Test		0.078	04:12
Air Blank		0.000	04:13
Diagnostics Check		OK	04:13

Data Download Date/Time: 04/06/2016 19:43

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEAVER,D,
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: COLSON,B
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 03/09/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A3
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 19:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:56

SUBJECT'S LAST NAME: YAHNKE

DATE/TIME OF TEST: 04/02/2016 21:26:04

Diagnostics Check		OK	21:19	
Air Blank		0.000	21:20	
Control Test		0.078	21:20	
Air Blank		0.000	21:20	
Subject Sample	# 1	0.268	21:21	
Breath Volume		1.703	21:21	
Air Blank		0.000	21:22	
Air Blank		0.000	21:23	
Subject Sample	# 2	0.275	21:24	
Breath Volume		1.328	21:24	
Air Blank		0.000	21:25	
Control Test		0.074	21:25	Control Outside Tolerance
Air Blank		0.000	21:26	

Data Download Date/Time: 04/29/2016 15:03

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEAVER,D,
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: COLSON,B
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 03/09/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A3
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 19:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:56

SUBJECT'S LAST NAME: YAHNKE

DATE/TIME OF TEST: 04/02/2016 21:34:12

Diagnostics Check		OK	21:27
Air Blank		0.000	21:28
Control Test		0.077	21:28
Air Blank		0.000	21:29
Subject Sample	# 1	0.271	21:29
Breath Volume		1.468	21:29
Air Blank		0.000	21:30
Air Blank		0.000	21:32
Subject Sample	# 2	0.289	21:32
Breath Volume		1.511	21:32
Air Blank		0.000	21:33
Control Test		0.077	21:33
Air Blank		0.000	21:34
Diagnostics Check		OK	21:34

Data Download Date/Time: 04/29/2016 15:03

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BEAUSOLEIL,AARON,F
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: BEAUSOLEIL,AARON
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 04/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A3
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:02

SUBJECT'S LAST NAME: OWENS

DATE/TIME OF TEST: 04/30/2016 23:28:48

Diagnostics Check		OK	23:22
Air Blank		0.000	23:23
Control Test		0.078	23:23
Air Blank		0.000	23:24
Subject Sample	# 1	0.110	23:24
Breath Volume		1.531	23:24
Air Blank		0.000	23:25
Air Blank		0.000	23:26
Subject Sample	# 2	0.107	23:27
Breath Volume		1.562	23:27
Air Blank		0.000	23:27
Control Test		0.079	23:28
Air Blank		0.000	23:28
Diagnostics Check		OK	23:28

Data Download Date/Time: 05/06/2016 12:39

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.