

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000223

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PULLEY,PATRICK,S
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: PULLEY,PATRICK
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 06/02/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:14

SUBJECT'S LAST NAME: MOTT

DATE/TIME OF TEST: 07/02/2016 01:41:21

Diagnostics Check	OK	01:34
Air Blank	0.000	01:35
Control Test	0.079	01:35
Air Blank	0.000	01:36
Subject Sample # 1	0.222	01:36
Breath Volume	2.378	01:36
Air Blank	0.000	01:37
Air Blank	0.000	01:39
Subject Sample # 2	0.221	01:39
Breath Volume	2.703	01:39
Air Blank	0.000	01:40
Control Test	0.077	01:40
Air Blank	0.000	01:41
Diagnostics Check	OK	01:41

Data Download Date/Time: 07/07/2016 11:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000223

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOFF,SARA,B
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: GOFF,SARA
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 06/02/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: MEJIA REYES

DATE/TIME OF TEST: 07/02/2016 23:56:54

Diagnostics Check	OK	23:50
Air Blank	0.000	23:51
Control Test	0.078	23:51
Air Blank	0.000	23:52
Subject Sample # 1	0.151	23:52
Breath Volume	2.394	23:52
Air Blank	0.000	23:53
Air Blank	0.000	23:54
Subject Sample # 2	0.156	23:55
Breath Volume	1.636	23:55
Air Blank	0.000	23:55
Control Test	0.078	23:56
Air Blank	0.000	23:56
Diagnostics Check	OK	23:56

Data Download Date/Time: 07/07/2016 11:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000223

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCAPIN,BRANDON,L
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: SCAPIN,BRANDON
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 07/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:53

SUBJECT'S LAST NAME: CARRUTH

DATE/TIME OF TEST: 07/28/2016 00:20:44

Diagnostics Check		OK	00:14
Air Blank		0.000	00:14
Control Test		0.077	00:14
Air Blank		0.000	00:15
Subject Sample	# 1	0.224	00:15
Breath Volume		1.667	00:15
Air Blank		0.000	00:16
Air Blank		0.000	00:18
Subject Sample	# 2	0.232	00:19
Breath Volume		1.433	00:19
Air Blank		0.000	00:19
Control Test		0.077	00:20
Air Blank		0.000	00:20
Diagnostics Check		OK	00:20

Data Download Date/Time: 08/27/2016 21:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRICKER,JEFFREY,S
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: FRICKER,JEFFREY
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 06/27/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 08:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:00

SUBJECT'S LAST NAME: LUCIANI

DATE/TIME OF TEST: 07/02/2016 09:55:25

Diagnostics Check	OK	09:47	
Air Blank	0.000	09:47	
Control Test	0.078	09:47	
Air Blank	0.000	09:48	
Subject Sample	# 1	09:51	No Sample Provided
Breath Volume	0.000	09:51	
Air Blank	0.000	09:51	
Air Blank	0.000	09:53	
Subject Sample	# 2	09:54	Subject Test Refused
Breath Volume	0.000	09:54	
Air Blank	0.000	09:54	
Control Test	0.078	09:54	
Air Blank	0.000	09:55	
Diagnostics Check	OK	09:55	

Data Download Date/Time: 07/26/2016 09:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUNT,CHARLES,R
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: HUNT,CHARLES
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 06/27/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: MITCHELL

DATE/TIME OF TEST: 07/04/2016 03:51:13

Diagnostics Check	OK	03:39	
Air Blank	0.000	03:40	
Control Test	0.078	03:40	
Air Blank	0.000	03:41	
Subject Sample	# 1	03:44	No Sample Provided
Breath Volume	0.000	03:44	
Air Blank	0.000	03:44	
Air Blank	0.000	03:46	
Subject Sample	# 2	03:49	No Sample Provided
Breath Volume	0.000	03:49	
Air Blank	0.000	03:50	
Control Test	0.079	03:50	
Air Blank	0.000	03:51	
Diagnostics Check	OK	03:51	

Data Download Date/Time: 07/26/2016 09:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUNT,CHARLES,R
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: HUNT,CHARLES
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 06/27/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:42

SUBJECT'S LAST NAME: DRAKE

DATE/TIME OF TEST: 07/08/2016 05:20:09

Diagnostics Check	OK	05:14
Air Blank	0.000	05:14
Control Test	0.078	05:14
Air Blank	0.000	05:15
Subject Sample # 1	0.216	05:15
Breath Volume	2.445	05:15
Air Blank	0.000	05:16
Air Blank	0.000	05:18
Subject Sample # 2	0.212	05:18
Breath Volume	1.667	05:18
Air Blank	0.000	05:19
Control Test	0.076	05:19
Air Blank	0.000	05:20
Diagnostics Check	OK	05:20

Data Download Date/Time: 07/26/2016 09:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUNT,CHARLES,R
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: HUNT,CHARLES
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 06/27/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:21

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 07/10/2016 01:58:38

Diagnostics Check	OK	01:52
Air Blank	0.000	01:52
Control Test	0.080	01:52
Air Blank	0.000	01:53
Subject Sample # 1	0.000	01:54
Breath Volume	1.992	01:54
Air Blank	0.000	01:54
Air Blank	0.000	01:56
Subject Sample # 2	0.000	01:57
Breath Volume	1.914	01:57
Air Blank	0.000	01:57
Control Test	0.077	01:58
Air Blank	0.000	01:58
Diagnostics Check	OK	01:58

Data Download Date/Time: 07/26/2016 09:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUNT,CHARLES,R
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: IANNUCCI,NICOLAS
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 06/27/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:42

SUBJECT'S LAST NAME: MINOR

DATE/TIME OF TEST: 07/10/2016 03:14:18

Diagnostics Check	OK	03:02	
Air Blank	0.000	03:03	
Control Test	0.078	03:03	
Air Blank	0.000	03:04	
Subject Sample	# 1	03:07	No Sample Provided
Breath Volume	0.000	03:07	
Air Blank	0.000	03:07	
Air Blank	0.000	03:09	
Subject Sample	# 2	03:12	No Sample Provided
Breath Volume	0.000	03:12	
Air Blank	0.000	03:13	
Control Test	0.079	03:13	
Air Blank	0.000	03:14	
Diagnostics Check	OK	03:14	

Data Download Date/Time: 07/26/2016 09:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNS, WILLIAM, A
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: PINO-DIAZ, VICTOR
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 06/27/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: MCCRAY

DATE/TIME OF TEST: 07/10/2016 04:18:51

Diagnostics Check	OK	04:12
Air Blank	0.000	04:13
Control Test	0.079	04:13
Air Blank	0.000	04:13
Subject Sample # 1	0.184	04:14
Breath Volume	1.585	04:14
Air Blank	0.000	04:15
Air Blank	0.000	04:16
Subject Sample # 2	0.180	04:17
Breath Volume	1.609	04:17
Air Blank	0.000	04:17
Control Test	0.076	04:18
Air Blank	0.000	04:18
Diagnostics Check	OK	04:18

Data Download Date/Time: 07/26/2016 09:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUNT,CHARLES,R
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: HUNT,CHARLES
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 06/27/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:36

SUBJECT'S LAST NAME: GEORGE-CARTER

DATE/TIME OF TEST: 07/10/2016 05:03:58

Diagnostics Check	OK	04:57
Air Blank	0.000	04:58
Control Test	0.079	04:58
Air Blank	0.000	04:59
Subject Sample # 1	0.056	04:59
Breath Volume	3.761	04:59
Air Blank	0.000	05:00
Air Blank	0.000	05:02
Subject Sample # 2	0.050	05:02
Breath Volume	2.468	05:02
Air Blank	0.000	05:03
Control Test	0.077	05:03
Air Blank	0.000	05:03
Diagnostics Check	OK	05:03

Data Download Date/Time: 07/26/2016 09:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOSTICK,JESSE,W
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: LAKE,JAMES
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 06/27/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:18

SUBJECT'S LAST NAME: BENTLEY

DATE/TIME OF TEST: 07/12/2016 03:47:15

Diagnostics Check	OK	03:39
Air Blank	0.000	03:39
Control Test	0.078	03:40
Air Blank	0.000	03:40
Subject Sample # 1	0.189	03:42
Breath Volume	1.253	03:42
Air Blank	0.000	03:43
Air Blank	0.000	03:44
Subject Sample # 2	0.184	03:45
Breath Volume	1.527	03:45
Air Blank	0.000	03:46
Control Test	0.075	03:46
Air Blank	0.000	03:47
Diagnostics Check	OK	03:47

Data Download Date/Time: 07/26/2016 09:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUNT,CHARLES,R
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: HUNT,CHARLES
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 06/27/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:13

SUBJECT'S LAST NAME: FULTZ

DATE/TIME OF TEST: 07/13/2016 04:54:57

Diagnostics Check	OK	04:43	
Air Blank	0.000	04:44	
Control Test	0.077	04:44	
Air Blank	0.000	04:44	
Subject Sample	# 1	04:47	No Sample Provided
Breath Volume	0.000	04:47	
Air Blank	0.000	04:48	
Air Blank	0.000	04:50	
Subject Sample	# 2	04:53	No Sample Provided
Breath Volume	0.000	04:53	
Air Blank	0.000	04:54	
Control Test	0.079	04:54	
Air Blank	0.000	04:54	
Diagnostics Check	OK	04:54	

Data Download Date/Time: 07/26/2016 09:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICHARDSON, MICHAEL, R
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: BOHANNON, MICHAEL
ARRESTING OFFICER AGENCY: Alachua PD
LAST AGENCY INSPECTION: 06/27/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:18

SUBJECT'S LAST NAME: FOWLER

DATE/TIME OF TEST: 07/14/2016 03:05:14

Diagnostics Check	OK	02:53	
Air Blank	0.000	02:54	
Control Test	0.078	02:54	
Air Blank	0.000	02:55	
Subject Sample	# 1	02:58	No Sample Provided
Breath Volume	0.000	02:58	
Air Blank	0.000	02:58	
Air Blank	0.000	03:00	
Subject Sample	# 2	03:03	No Sample Provided
Breath Volume	0.000	03:03	
Air Blank	0.000	03:04	
Control Test	0.080	03:04	
Air Blank	0.000	03:05	
Diagnostics Check	OK	03:05	

Data Download Date/Time: 07/26/2016 09:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUNT,CHARLES,R
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: HUNT,CHARLES
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 06/27/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:56

SUBJECT'S LAST NAME: MCCRAY

DATE/TIME OF TEST: 07/16/2016 04:34:27

Diagnostics Check	OK	04:28
Air Blank	0.000	04:28
Control Test	0.078	04:29
Air Blank	0.000	04:29
Subject Sample # 1	0.163	04:30
Breath Volume	2.933	04:30
Air Blank	0.000	04:30
Air Blank	0.000	04:32
Subject Sample # 2	0.171	04:32
Breath Volume	1.847	04:32
Air Blank	0.000	04:33
Control Test	0.077	04:33
Air Blank	0.000	04:34
Diagnostics Check	OK	04:34

Data Download Date/Time: 07/26/2016 09:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUNT,CHARLES,R
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: HUNT,CHARLES
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 06/27/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:44

SUBJECT'S LAST NAME: WILLIS

DATE/TIME OF TEST: 07/16/2016 21:23:57

Diagnostics Check	OK	21:14	
Air Blank	0.000	21:14	
Control Test	0.078	21:15	
Air Blank	0.000	21:15	
Subject Sample	# 1 0.333	21:16	
Breath Volume	2.304	21:16	
Air Blank	0.000	21:16	
Air Blank	0.000	21:18	
Subject Sample	# 2 0.293	21:19	
Breath Volume	1.332	21:19	
Air Blank	0.000	21:20	
Air Blank	0.000	21:21	
Subject Sample	# 3 0.315	21:22	
Breath Volume	2.898	21:22	
Air Blank	0.000	21:23	
Control Test	0.074	21:23	Control Outside Tolerance
Air Blank	0.000	21:23	

Data Download Date/Time: 07/26/2016 09:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOFMANN,TRAVIS,A
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: HOFMANN,TRAVIS
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 06/27/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: JAYNE

DATE/TIME OF TEST: 07/17/2016 00:30:55

Diagnostics Check	OK	00:24
Air Blank	0.000	00:25
Control Test	0.077	00:25
Air Blank	0.000	00:25
Subject Sample # 1	0.190	00:26
Breath Volume	1.757	00:26
Air Blank	0.000	00:26
Air Blank	0.000	00:28
Subject Sample # 2	0.196	00:29
Breath Volume	1.578	00:29
Air Blank	0.000	00:30
Control Test	0.079	00:30
Air Blank	0.000	00:30
Diagnostics Check	OK	00:30

Data Download Date/Time: 07/26/2016 09:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,JACK,W
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: MILLER,JACK
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 06/27/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 07:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:44

SUBJECT'S LAST NAME: WEAVER

DATE/TIME OF TEST: 07/17/2016 08:25:01

Diagnostics Check	OK	08:18
Air Blank	0.000	08:19
Control Test	0.078	08:19
Air Blank	0.000	08:20
Subject Sample # 1	0.000	08:20
Breath Volume	2.773	08:20
Air Blank	0.000	08:21
Air Blank	0.000	08:23
Subject Sample # 2	0.000	08:23
Breath Volume	3.261	08:23
Air Blank	0.000	08:24
Control Test	0.077	08:24
Air Blank	0.000	08:24
Diagnostics Check	OK	08:25

Data Download Date/Time: 07/26/2016 09:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOSTICK,MEGAN,K
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: SINGLETARY,KENNETH
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 06/27/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 17:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:50

SUBJECT'S LAST NAME: CANNIZZO

DATE/TIME OF TEST: 07/18/2016 18:28:15

Diagnostics Check	OK	18:21
Air Blank	0.000	18:21
Control Test	0.078	18:22
Air Blank	0.000	18:22
Subject Sample # 1	0.332	18:23
Breath Volume	3.378	18:23
Air Blank	0.000	18:24
Air Blank	0.000	18:26
Subject Sample # 2	0.340	18:26
Breath Volume	1.683	18:26
Air Blank	0.000	18:27
Control Test	0.077	18:27
Air Blank	0.000	18:28
Diagnostics Check	OK	18:28

Data Download Date/Time: 07/26/2016 09:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUNT,CHARLES,R
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: HUNT,CHARLES
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 06/27/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:24

SUBJECT'S LAST NAME: GRIMM

DATE/TIME OF TEST: 07/20/2016 00:59:52

Diagnostics Check	OK	00:53
Air Blank	0.000	00:54
Control Test	0.078	00:54
Air Blank	0.000	00:54
Subject Sample # 1	0.127	00:55
Breath Volume	2.328	00:55
Air Blank	0.000	00:56
Air Blank	0.000	00:57
Subject Sample # 2	0.126	00:58
Breath Volume	2.269	00:58
Air Blank	0.000	00:58
Control Test	0.077	00:59
Air Blank	0.000	00:59
Diagnostics Check	OK	00:59

Data Download Date/Time: 07/26/2016 09:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUALMANN,DANIEL,G
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: QUALMANN,DANIEL
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 06/27/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:12

SUBJECT'S LAST NAME: DEES

DATE/TIME OF TEST: 07/23/2016 03:45:21

Diagnostics Check	OK	03:39
Air Blank	0.000	03:39
Control Test	0.078	03:39
Air Blank	0.000	03:40
Subject Sample # 1	0.113	03:40
Breath Volume	3.687	03:40
Air Blank	0.000	03:41
Air Blank	0.000	03:43
Subject Sample # 2	0.103	03:43
Breath Volume	3.726	03:43
Air Blank	0.000	03:44
Control Test	0.077	03:44
Air Blank	0.000	03:45
Diagnostics Check	OK	03:45

Data Download Date/Time: 07/26/2016 09:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUALMANN,DANIEL,G
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: QUALMANN,DANIEL
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 06/27/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:43

SUBJECT'S LAST NAME: PEREZ-ROBLERO

DATE/TIME OF TEST: 07/24/2016 04:30:15

Diagnostics Check	OK	04:23
Air Blank	0.000	04:24
Control Test	0.078	04:24
Air Blank	0.000	04:24
Subject Sample # 1	0.225	04:25
Breath Volume	3.257	04:25
Air Blank	0.000	04:26
Air Blank	0.000	04:28
Subject Sample # 2	0.208	04:28
Breath Volume	1.871	04:28
Air Blank	0.000	04:29
Control Test	0.076	04:29
Air Blank	0.000	04:30
Diagnostics Check	OK	04:30

Data Download Date/Time: 07/26/2016 09:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LARA,ALEXANDER,
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: LARA,ALEXANDER
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 07/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: CANNON

DATE/TIME OF TEST: 07/27/2016 02:03:46

Diagnostics Check	OK	01:55	
Air Blank	0.000	01:55	
Control Test	0.078	01:56	
Air Blank	0.000	01:56	
Subject Sample	# 1	01:59	No Sample Provided
Breath Volume	0.000	01:59	
Air Blank	0.000	02:00	
Air Blank	0.000	02:02	
Subject Sample	# 2	02:02	Subject Test Refused
Breath Volume	0.000	02:02	
Air Blank	0.000	02:02	
Control Test	0.079	02:03	
Air Blank	0.000	02:03	
Diagnostics Check	OK	02:03	

Data Download Date/Time: 08/25/2016 10:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUNT,CHARLES,R
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: HUNT,CHARLES
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 07/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:43

SUBJECT'S LAST NAME: BAUTEL

DATE/TIME OF TEST: 07/28/2016 00:17:22

Diagnostics Check	OK	00:11
Air Blank	0.000	00:11
Control Test	0.079	00:11
Air Blank	0.000	00:12
Subject Sample # 1	0.000	00:13
Breath Volume	2.070	00:13
Air Blank	0.000	00:13
Air Blank	0.000	00:15
Subject Sample # 2	0.000	00:16
Breath Volume	1.996	00:16
Air Blank	0.000	00:16
Control Test	0.077	00:16
Air Blank	0.000	00:17
Diagnostics Check	OK	00:17

Data Download Date/Time: 08/25/2016 10:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUNT,CHARLES,R
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: HUNT,CHARLES
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 07/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 19:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:22

SUBJECT'S LAST NAME: FERGUSON

DATE/TIME OF TEST: 07/29/2016 20:57:00

Diagnostics Check	OK	20:50
Air Blank	0.000	20:51
Control Test	0.078	20:51
Air Blank	0.000	20:52
Subject Sample # 1	0.162	20:52
Breath Volume	1.746	20:52
Air Blank	0.000	20:53
Air Blank	0.000	20:55
Subject Sample # 2	0.160	20:55
Breath Volume	1.988	20:55
Air Blank	0.000	20:56
Control Test	0.076	20:56
Air Blank	0.000	20:56
Diagnostics Check	OK	20:57

Data Download Date/Time: 08/25/2016 10:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUNT,CHARLES,R
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: HUNT,CHARLES
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 07/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 12:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: FINN

DATE/TIME OF TEST: 07/30/2016 01:41:50

Diagnostics Check	OK	01:35
Air Blank	0.000	01:36
Control Test	0.079	01:36
Air Blank	0.000	01:36
Subject Sample # 1	0.097	01:37
Breath Volume	2.574	01:37
Air Blank	0.000	01:38
Air Blank	0.000	01:39
Subject Sample # 2	0.099	01:40
Breath Volume	2.867	01:40
Air Blank	0.000	01:40
Control Test	0.077	01:41
Air Blank	0.000	01:41
Diagnostics Check	OK	01:41

Data Download Date/Time: 08/25/2016 10:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNS, WILLIAM, A
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: JOHNS, WILLIAM
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 07/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 07/30/2016 04:01:21

Diagnostics Check	OK	03:53
Air Blank	0.000	03:54
Control Test	0.079	03:54
Air Blank	0.000	03:55
Subject Sample # 1	0.088	03:56
Breath Volume	1.742	03:56
Air Blank	0.000	03:57
Air Blank	0.000	03:59
Subject Sample # 2	0.089	03:59
Breath Volume	1.542	03:59
Air Blank	0.000	04:00
Control Test	0.075	04:00
Air Blank	0.000	04:01
Diagnostics Check	OK	04:01

Data Download Date/Time: 08/25/2016 10:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000738

INSTRUMENT REGISTERED TO: FT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRUJILLO,ANTHONY,R
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: TRUJILLO,ANTHONY
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 06/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A1
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:33

SUBJECT'S LAST NAME: LANDMESSER

DATE/TIME OF TEST: 07/03/2016 02:00:28

Diagnostics Check	OK	01:54
Air Blank	0.000	01:54
Control Test	0.081	01:55
Air Blank	0.000	01:55
Subject Sample # 1	0.186	01:56
Breath Volume	3.878	01:56
Air Blank	0.000	01:56
Air Blank	0.000	01:58
Subject Sample # 2	0.178	01:58
Breath Volume	4.132	01:58
Air Blank	0.000	01:59
Control Test	0.080	01:59
Air Blank	0.000	02:00
Diagnostics Check	OK	02:00

Data Download Date/Time: 08/04/2016 19:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000738

INSTRUMENT REGISTERED TO: FT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRUJILLO,ANTHONY,R
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: TRUJILLO,ANTHONY
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 06/29/2016

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 18015080A1
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:45

SUBJECT'S LAST NAME: KELLEY

DATE/TIME OF TEST: 07/04/2016 05:09:03

Diagnostics Check	OK	05:06	
Air Blank	0.000	05:06	
Control Test	0.080	05:07	
Air Blank	0.000	05:07	
Subject Sample	# 1	05:07	Subject Test Refused
Breath Volume	0.000	05:07	
Air Blank	0.000	05:08	
Control Test	0.080	05:08	
Air Blank	0.000	05:08	
Diagnostics Check	OK	05:09	

Data Download Date/Time: 08/04/2016 19:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000738

INSTRUMENT REGISTERED TO: FT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANTIAGO,EDGARDO,
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: SANTIAGO,EDGARDO
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 06/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A1
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: ROSSI

DATE/TIME OF TEST: 07/08/2016 03:52:44

Diagnostics Check	OK	03:41	
Air Blank	0.000	03:41	
Control Test	0.081	03:41	
Air Blank	0.000	03:42	
Subject Sample	# 1	03:45	No Sample Provided
Breath Volume	0.000	03:45	
Air Blank	0.000	03:46	
Air Blank	0.000	03:48	
Subject Sample	# 2	0.181	03:48
Breath Volume	3.660	03:48	
Air Blank	0.000	03:49	
Air Blank	0.000	03:50	
Subject Sample	# 3	0.175	03:51
Breath Volume	2.363	03:51	
Air Blank	0.000	03:51	
Control Test	0.080	03:52	
Air Blank	0.000	03:52	
Diagnostics Check	OK	03:52	

Data Download Date/Time: 08/04/2016 19:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000738

INSTRUMENT REGISTERED TO: FT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOGAN, MICHAEL, J
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: HOGAN, MICHAEL
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 06/29/2016

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 18015080A1
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:31

SUBJECT'S LAST NAME: KING

DATE/TIME OF TEST: 07/16/2016 02:01:21

Diagnostics Check	OK	01:58	
Air Blank	0.000	01:58	
Control Test	0.080	01:58	
Air Blank	0.000	01:59	
Subject Sample	# 1	02:00	Subject Test Refused
Breath Volume	0.000	02:00	
Air Blank	0.000	02:00	
Control Test	0.079	02:00	
Air Blank	0.000	02:01	
Diagnostics Check	OK	02:01	

Data Download Date/Time: 08/04/2016 19:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000738

INSTRUMENT REGISTERED TO: FT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSENBURG,DUSTIN,L
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: ROSENBURG,DUSTIN
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 06/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A1
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:38

SUBJECT'S LAST NAME: LONGMAN

DATE/TIME OF TEST: 07/17/2016 04:02:51

Diagnostics Check	OK	03:59	
Air Blank	0.000	04:00	
Control Test	0.081	04:00	
Air Blank	0.000	04:01	
Subject Sample	# 1	04:01	Subject Test Refused
Breath Volume	0.000	04:01	
Air Blank	0.000	04:01	
Control Test	0.079	04:02	
Air Blank	0.000	04:02	
Diagnostics Check	OK	04:02	

Data Download Date/Time: 08/04/2016 19:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000738

INSTRUMENT REGISTERED TO: FT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANTIAGO,EDGARDO,
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: SANTIAGO,EDGARDO
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 06/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A1
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 05:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:58

SUBJECT'S LAST NAME: FOSHEE

DATE/TIME OF TEST: 07/17/2016 06:28:44

Diagnostics Check	OK	06:19
Air Blank	0.000	06:20
Control Test	0.080	06:20
Air Blank	0.000	06:21
Subject Sample # 1	0.195	06:24
Breath Volume	1.882	06:24
Air Blank	0.000	06:24
Air Blank	0.000	06:26
Subject Sample # 2	0.201	06:27
Breath Volume	1.203	06:27
Air Blank	0.000	06:27
Control Test	0.079	06:28
Air Blank	0.000	06:28
Diagnostics Check	OK	06:28

Data Download Date/Time: 08/04/2016 19:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000738

INSTRUMENT REGISTERED TO: FT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHISHOLM,RYAN,D
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: CHISHOLM,RYAN
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 06/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A1
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 12:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:09

SUBJECT'S LAST NAME: SPENCE

DATE/TIME OF TEST: 07/17/2016 13:35:07

Diagnostics Check	OK	13:32	
Air Blank	0.000	13:32	
Control Test	0.079	13:32	
Air Blank	0.000	13:33	
Subject Sample	# 1	13:33	Subject Test Refused
Breath Volume	0.000	13:33	
Air Blank	0.000	13:34	
Control Test	0.079	13:34	
Air Blank	0.000	13:35	
Diagnostics Check	OK	13:35	

Data Download Date/Time: 08/04/2016 19:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000738

INSTRUMENT REGISTERED TO: FT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESNEUL,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: ESNEUL,MICHAEL
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 07/31/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A1
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:14

SUBJECT'S LAST NAME: AMADOR

DATE/TIME OF TEST: 07/31/2016 23:45:30

Diagnostics Check	OK	23:36	
Air Blank	0.000	23:37	
Control Test	0.079	23:37	
Air Blank	0.000	23:38	
Subject Sample	# 1	23:41	No Sample Provided
Breath Volume	0.000	23:41	
Air Blank	0.000	23:41	
Air Blank	0.000	23:43	
Subject Sample	# 2	23:44	Subject Test Refused
Breath Volume	0.000	23:44	
Air Blank	0.000	23:44	
Control Test	0.080	23:44	
Air Blank	0.000	23:45	
Diagnostics Check	OK	23:45	

Data Download Date/Time: 08/04/2016 19:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GENRICH,KRISTEN,R
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: HIPWELL,PATRICIA
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 06/27/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 06:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:10

SUBJECT'S LAST NAME: CARROLL

DATE/TIME OF TEST: 07/02/2016 07:37:57

Diagnostics Check	OK	07:30
Air Blank	0.000	07:31
Control Test	0.078	07:31
Air Blank	0.000	07:31
Subject Sample # 1	0.170	07:33
Breath Volume	3.230	07:33
Air Blank	0.000	07:33
Air Blank	0.000	07:35
Subject Sample # 2	0.166	07:36
Breath Volume	2.101	07:36
Air Blank	0.000	07:37
Control Test	0.077	07:37
Air Blank	0.000	07:37
Diagnostics Check	OK	07:37

Data Download Date/Time: 07/26/2016 12:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHRISTMAS, MATTHEW, S
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: MILONAS, JO
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 06/27/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 13:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:34

SUBJECT'S LAST NAME: CLECKER

DATE/TIME OF TEST: 07/02/2016 14:20:56

Diagnostics Check	OK	14:12
Air Blank	0.000	14:12
Control Test	0.077	14:13
Air Blank	0.000	14:13
Subject Sample # 1	0.238	14:14
Breath Volume	2.218	14:14
Air Blank	0.000	14:15
Air Blank	0.000	14:17
Subject Sample # 2	0.235	14:19
Breath Volume	2.644	14:19
Air Blank	0.000	14:20
Control Test	0.076	14:20
Air Blank	0.000	14:20
Diagnostics Check	OK	14:20

Data Download Date/Time: 07/26/2016 12:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KEEL,JESSICA,M
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: KEEL,JESSICA
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 06/27/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 07/05/2016 02:18:23

Diagnostics Check	OK	02:07	
Air Blank	0.000	02:07	
Control Test	0.080	02:07	
Air Blank	0.000	02:08	
Subject Sample	# 1	02:11	No Sample Provided
Breath Volume	0.000	02:11	
Air Blank	0.000	02:11	
Air Blank	0.000	02:13	
Subject Sample	# 2	02:17	No Sample Provided
Breath Volume	0.000	02:17	
Air Blank	0.000	02:17	
Control Test	0.079	02:17	
Air Blank	0.000	02:18	
Diagnostics Check	OK	02:18	

Data Download Date/Time: 07/26/2016 12:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNOWLES, CONSTANCE, J
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: KNOWLES, CONSTANCE
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 06/27/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:24

SUBJECT'S LAST NAME: PENROD

DATE/TIME OF TEST: 07/16/2016 02:59:07

Diagnostics Check	OK	02:52
Air Blank	0.000	02:53
Control Test	0.080	02:53
Air Blank	0.000	02:54
Subject Sample # 1	0.081	02:54
Breath Volume	3.730	02:54
Air Blank	0.000	02:55
Air Blank	0.000	02:57
Subject Sample # 2	0.085	02:57
Breath Volume	3.976	02:57
Air Blank	0.000	02:58
Control Test	0.079	02:58
Air Blank	0.000	02:59
Diagnostics Check	OK	02:59

Data Download Date/Time: 07/26/2016 12:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNOWLES, CONSTANCE, J
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: KNOWLES, CONSTANCE
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 06/27/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 06:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:01

SUBJECT'S LAST NAME: TARPLEY

DATE/TIME OF TEST: 07/16/2016 07:54:14

Diagnostics Check	OK	07:51	
Air Blank	0.000	07:51	
Control Test	0.078	07:52	
Air Blank	0.000	07:52	
Subject Sample	# 1	07:52	Subject Test Refused
Breath Volume	0.000	07:52	
Air Blank	0.000	07:53	
Control Test	0.077	07:53	
Air Blank	0.000	07:54	
Diagnostics Check	OK	07:54	

Data Download Date/Time: 07/26/2016 12:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBINSON,ROBERT,S
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: DOWDY,JEREMIAH
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 06/27/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 11:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 11:45

SUBJECT'S LAST NAME: MATHEWS

DATE/TIME OF TEST: 07/17/2016 12:14:01

Diagnostics Check	OK	12:07
Air Blank	0.000	12:07
Control Test	0.077	12:08
Air Blank	0.000	12:08
Subject Sample # 1	0.000	12:09
Breath Volume	2.984	12:09
Air Blank	0.000	12:10
Air Blank	0.000	12:12
Subject Sample # 2	0.000	12:12
Breath Volume	2.070	12:12
Air Blank	0.000	12:13
Control Test	0.078	12:13
Air Blank	0.000	12:13
Diagnostics Check	OK	12:14

Data Download Date/Time: 07/26/2016 12:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANE,AMANDA,M
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: LANE,AMANDA
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 06/27/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:00

SUBJECT'S LAST NAME: WAITS

DATE/TIME OF TEST: 07/19/2016 05:31:10

Diagnostics Check	OK	05:24
Air Blank	0.000	05:24
Control Test	0.078	05:25
Air Blank	0.000	05:25
Subject Sample # 1	0.144	05:26
Breath Volume	2.406	05:26
Air Blank	0.000	05:27
Air Blank	0.000	05:28
Subject Sample # 2	0.147	05:29
Breath Volume	3.597	05:29
Air Blank	0.000	05:30
Control Test	0.077	05:30
Air Blank	0.000	05:31
Diagnostics Check	OK	05:31

Data Download Date/Time: 07/26/2016 12:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANE,AMANDA,M
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: LANE,AMANDA
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 06/27/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:49

SUBJECT'S LAST NAME: GRIGGS

DATE/TIME OF TEST: 07/24/2016 03:16:53

Diagnostics Check	OK	03:10
Air Blank	0.000	03:11
Control Test	0.078	03:11
Air Blank	0.000	03:12
Subject Sample # 1	0.000	03:12
Breath Volume	1.441	03:12
Air Blank	0.000	03:12
Air Blank	0.000	03:15
Subject Sample # 2	0.000	03:15
Breath Volume	2.554	03:15
Air Blank	0.000	03:15
Control Test	0.077	03:16
Air Blank	0.000	03:16
Diagnostics Check	OK	03:16

Data Download Date/Time: 07/26/2016 12:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANE,AMANDA,M
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: LANE,AMANDA
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 06/27/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:49

SUBJECT'S LAST NAME: GRIGGS

DATE/TIME OF TEST: 07/24/2016 03:25:41

Diagnostics Check	OK	03:19
Air Blank	0.000	03:19
Control Test	0.078	03:20
Air Blank	0.000	03:20
Subject Sample # 1	0.000	03:21
Breath Volume	2.164	03:21
Air Blank	0.000	03:21
Air Blank	0.000	03:23
Subject Sample # 2	0.000	03:24
Breath Volume	2.242	03:24
Air Blank	0.000	03:24
Control Test	0.077	03:25
Air Blank	0.000	03:25
Diagnostics Check	OK	03:25

Data Download Date/Time: 07/26/2016 12:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREIRA,JOSE,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: PEREIRA,JOSE
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 06/27/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:14

SUBJECT'S LAST NAME: PUTNEY

DATE/TIME OF TEST: 07/24/2016 04:50:52

Diagnostics Check	OK	04:42	
Air Blank	0.000	04:42	
Control Test	0.079	04:42	
Air Blank	0.000	04:43	
Subject Sample	# 1	04:46	No Sample Provided
Breath Volume	0.000	04:46	
Air Blank	0.000	04:46	
Air Blank	0.000	04:48	
Subject Sample	# 2	04:49	Subject Test Refused
Breath Volume	0.000	04:49	
Air Blank	0.000	04:49	
Control Test	0.078	04:50	
Air Blank	0.000	04:50	
Diagnostics Check	OK	04:50	

Data Download Date/Time: 07/26/2016 12:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREIRA,JOSE,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: EMRICH,IAN
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 06/27/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:55

SUBJECT'S LAST NAME: THOMPSON JR

DATE/TIME OF TEST: 07/24/2016 05:26:08

Diagnostics Check	OK	05:18
Air Blank	0.000	05:18
Control Test	0.078	05:18
Air Blank	0.000	05:19
Subject Sample # 1	0.173	05:21
Breath Volume	1.929	05:21
Air Blank	0.000	05:21
Air Blank	0.000	05:23
Subject Sample # 2	0.161	05:24
Breath Volume	1.593	05:24
Air Blank	0.000	05:25
Control Test	0.077	05:25
Air Blank	0.000	05:26
Diagnostics Check	OK	05:26

Data Download Date/Time: 07/26/2016 12:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREIRA,JOSE,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: BOWEN,NATHANIEL
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 06/27/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:35

SUBJECT'S LAST NAME: MACARIO LOPEZ

DATE/TIME OF TEST: 07/24/2016 06:02:51

Diagnostics Check		OK	05:55
Air Blank		0.000	05:56
Control Test		0.077	05:56
Air Blank		0.000	05:57
Subject Sample	# 1	0.212	05:58
Breath Volume		2.296	05:58
Air Blank		0.000	05:58
Air Blank		0.000	06:00
Subject Sample	# 2	0.215	06:01
Breath Volume		2.468	06:01
Air Blank		0.000	06:01
Control Test		0.077	06:02
Air Blank		0.000	06:02
Diagnostics Check		OK	06:02

Data Download Date/Time: 07/26/2016 12:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KEEL,JESSICA,M
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: DEGENNARO,BRIANA
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 06/27/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: VICKERY

DATE/TIME OF TEST: 07/25/2016 02:27:30

Diagnostics Check	OK	02:18	
Air Blank	0.000	02:19	
Control Test	0.077	02:19	
Air Blank	0.000	02:20	
Subject Sample	# 1	02:23	No Sample Provided
Breath Volume	0.000	02:23	
Air Blank	0.000	02:23	
Air Blank	0.000	02:25	
Subject Sample	# 2	02:26	Subject Test Refused
Breath Volume	0.000	02:26	
Air Blank	0.000	02:26	
Control Test	0.077	02:26	
Air Blank	0.000	02:27	
Diagnostics Check	OK	02:27	

Data Download Date/Time: 07/26/2016 12:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KEEL,JESSICA,M
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: DEGENNARO,BRIANA
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 06/27/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: MESINER

DATE/TIME OF TEST: 07/25/2016 02:56:49

Diagnostics Check	OK	02:50
Air Blank	0.000	02:50
Control Test	0.077	02:51
Air Blank	0.000	02:51
Subject Sample # 1	0.160	02:52
Breath Volume	2.281	02:52
Air Blank	0.000	02:53
Air Blank	0.000	02:54
Subject Sample # 2	0.161	02:55
Breath Volume	2.050	02:55
Air Blank	0.000	02:55
Control Test	0.076	02:56
Air Blank	0.000	02:56
Diagnostics Check	OK	02:56

Data Download Date/Time: 07/26/2016 12:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SARMENT,JOSEPH,D
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: DEGENNARO,BRIANA
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 07/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:09

SUBJECT'S LAST NAME: WELSH

DATE/TIME OF TEST: 07/28/2016 01:32:59

Diagnostics Check	OK	01:30	
Air Blank	0.000	01:30	
Control Test	0.076	01:30	
Air Blank	0.000	01:31	
Subject Sample	# 1	01:31	Subject Test Refused
Breath Volume	0.000	01:31	
Air Blank	0.000	01:32	
Control Test	0.076	01:32	
Air Blank	0.000	01:32	
Diagnostics Check	OK	01:32	

Data Download Date/Time: 08/31/2016 11:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEVENS, MITCHELL, M
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: STEVENS, MITCHELL
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 06/27/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:38

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 07/01/2016 00:39:21

Diagnostics Check	OK	00:31
Air Blank	0.000	00:31
Control Test	0.078	00:32
Air Blank	0.000	00:32
Subject Sample # 1	0.174	00:34
Breath Volume	1.472	00:34
Air Blank	0.000	00:35
Air Blank	0.000	00:37
Subject Sample # 2	0.168	00:37
Breath Volume	1.457	00:37
Air Blank	0.000	00:38
Control Test	0.075	00:38
Air Blank	0.000	00:39
Diagnostics Check	OK	00:39

Data Download Date/Time: 07/12/2016 08:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRISSETT,ROMAIN,A
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: GRAVES,TIMOTHY
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 06/27/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 09:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:12

SUBJECT'S LAST NAME: KLEIN

DATE/TIME OF TEST: 07/02/2016 10:36:34

Diagnostics Check	OK	10:33	
Air Blank	0.000	10:33	
Control Test	0.078	10:33	
Air Blank	0.000	10:34	
Subject Sample	# 1	10:35	Subject Test Refused
Breath Volume	0.000	10:35	
Air Blank	0.000	10:35	
Control Test	0.078	10:35	
Air Blank	0.000	10:36	
Diagnostics Check	OK	10:36	

Data Download Date/Time: 07/12/2016 08:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAIRE,TRENTON,M
BREATH TEST OPERATOR AGENCY: OTHER Fed
ARRESTING OFFICER: HAIRE,TRENTON
ARRESTING OFFICER AGENCY: OTHER Fed
LAST AGENCY INSPECTION: 06/27/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: GREER

DATE/TIME OF TEST: 07/02/2016 23:09:09

Diagnostics Check	OK	23:04	
Air Blank	0.000	23:05	
Control Test	0.078	23:05	
Air Blank	0.000	23:06	
Subject Sample	# 1	23:08	Sequence Aborted
Breath Volume	0.097	23:08	
Air Blank	0.000	23:09	

Data Download Date/Time: 07/12/2016 08:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAIRE,TRENTON,M
BREATH TEST OPERATOR AGENCY: OTHER Fed
ARRESTING OFFICER: HAIRE,TRENTON
ARRESTING OFFICER AGENCY: OTHER Fed
LAST AGENCY INSPECTION: 06/27/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: GREER

DATE/TIME OF TEST: 07/02/2016 23:23:11

Diagnostics Check	OK	23:11		
Air Blank	0.000	23:11		
Control Test	0.078	23:12		
Air Blank	0.000	23:12		
Subject Sample	# 1	23:15	No Sample Provided	
Breath Volume	0.000	23:15		
Air Blank	0.000	23:16		
Air Blank	0.000	23:18		
Subject Sample	# 2	0.066	23:21	Volume Not Met
Breath Volume	0.730	23:21		
Air Blank	0.000	23:22		
Control Test	0.079	23:22		
Air Blank	0.000	23:23		
Diagnostics Check	OK	23:23		

Data Download Date/Time: 07/12/2016 08:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAYO,TREVOR,W
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: MAYO,TREVOR
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 06/27/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:54

SUBJECT'S LAST NAME: OBERHAUSEN

DATE/TIME OF TEST: 07/03/2016 03:37:03

Diagnostics Check	OK	03:28
Air Blank	0.000	03:28
Control Test	0.078	03:28
Air Blank	0.000	03:29
Subject Sample # 1	0.169	03:30
Breath Volume	2.769	03:30
Air Blank	0.000	03:30
Air Blank	0.000	03:32
Subject Sample # 2	0.170	03:35
Breath Volume	2.539	03:35
Air Blank	0.000	03:36
Control Test	0.078	03:36
Air Blank	0.000	03:36
Diagnostics Check	OK	03:37

Data Download Date/Time: 07/12/2016 08:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAYO,TREVOR,W
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: CRAMER,JONATHAN
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 07/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 18:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:00

SUBJECT'S LAST NAME: ARRANT

DATE/TIME OF TEST: 07/19/2016 20:44:15

Diagnostics Check	OK	20:37
Air Blank	0.000	20:38
Control Test	0.078	20:38
Air Blank	0.000	20:38
Subject Sample # 1	0.000	20:39
Breath Volume	2.363	20:39
Air Blank	0.000	20:39
Air Blank	0.000	20:42
Subject Sample # 2	0.000	20:42
Breath Volume	4.207	20:42
Air Blank	0.000	20:43
Control Test	0.077	20:43
Air Blank	0.000	20:44
Diagnostics Check	OK	20:44

Data Download Date/Time: 08/23/2016 10:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROOP,CHAD,E
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: ROOP,CHAD
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 07/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:47

SUBJECT'S LAST NAME: WEIR

DATE/TIME OF TEST: 07/19/2016 23:19:48

Diagnostics Check	OK	23:07
Air Blank	0.000	23:08
Control Test	0.077	23:08
Air Blank	0.000	23:09
Subject Sample # 1	0.185	23:11
Breath Volume	1.398	23:11
Air Blank	0.000	23:12
Air Blank	0.000	23:13
Subject Sample # 2	0.207	23:14
Breath Volume	1.378	23:14
Air Blank	0.000	23:15
Air Blank	0.000	23:17
Subject Sample # 3	0.200	23:18
Breath Volume	1.191	23:18
Air Blank	0.000	23:18
Control Test	0.077	23:19
Air Blank	0.000	23:19
Diagnostics Check	OK	23:19

Data Download Date/Time: 08/23/2016 10:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAIRE,TRENTON,M
BREATH TEST OPERATOR AGENCY: OTHER Fed
ARRESTING OFFICER: HAIRE,TRENTON
ARRESTING OFFICER AGENCY: OTHER Fed
LAST AGENCY INSPECTION: 07/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 12:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:22

SUBJECT'S LAST NAME: BARTLESON

DATE/TIME OF TEST: 07/23/2016 01:51:34

Diagnostics Check	OK	01:43
Air Blank	0.000	01:44
Control Test	0.079	01:44
Air Blank	0.000	01:45
Subject Sample # 1	0.126	01:46
Breath Volume	2.882	01:46
Air Blank	0.000	01:46
Air Blank	0.000	01:48
Subject Sample # 2	0.133	01:49
Breath Volume	1.796	01:49
Air Blank	0.000	01:50
Control Test	0.077	01:50
Air Blank	0.000	01:51
Diagnostics Check	OK	01:51

Data Download Date/Time: 08/23/2016 10:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROOP,CHAD,E
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: ROOP,CHAD
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 07/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:08

SUBJECT'S LAST NAME: TEACHEY

DATE/TIME OF TEST: 07/23/2016 22:29:50

Diagnostics Check	OK	22:28	
Air Blank	0.000	22:28	
Control Test	0.079	22:29	
Air Blank		22:29	RFI Detect
Air Blank	0.000	22:29	

Data Download Date/Time: 08/23/2016 10:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROOP,CHAD,E
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: ROOP,CHAD
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 07/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:08

SUBJECT'S LAST NAME: TEACHEY

DATE/TIME OF TEST: 07/23/2016 22:37:44

Diagnostics Check	OK	22:31
Air Blank	0.000	22:32
Control Test	0.079	22:32
Air Blank	0.000	22:32
Subject Sample # 1	0.000	22:33
Breath Volume	2.246	22:33
Air Blank	0.000	22:33
Air Blank	0.000	22:35
Subject Sample # 2	0.000	22:36
Breath Volume	1.578	22:36
Air Blank	0.000	22:36
Control Test	0.078	22:37
Air Blank	0.000	22:37
Diagnostics Check	OK	22:37

Data Download Date/Time: 08/23/2016 10:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEETERS,OZZIE,H
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: TEETERS,OZZIE
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 07/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:43

SUBJECT'S LAST NAME: HICKS

DATE/TIME OF TEST: 07/27/2016 23:17:52

Diagnostics Check	OK	23:06	
Air Blank	0.000	23:06	
Control Test	0.078	23:07	
Air Blank	0.000	23:07	
Subject Sample	# 1	23:10	No Sample Provided
Breath Volume	0.000	23:10	
Air Blank	0.000	23:11	
Air Blank	0.000	23:13	
Subject Sample	# 2	23:16	No Sample Provided
Breath Volume	0.000	23:16	
Air Blank	0.000	23:16	
Control Test	0.078	23:17	
Air Blank	0.000	23:17	
Diagnostics Check	OK	23:17	

Data Download Date/Time: 08/23/2016 10:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000759

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUIVEY,CYNTHIA,L
BREATH TEST OPERATOR AGENCY: Jackson County SO
ARRESTING OFFICER: WATSON,VERGIL
ARRESTING OFFICER AGENCY: Jackson County SO
LAST AGENCY INSPECTION: 06/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A3
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 14:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:58

SUBJECT'S LAST NAME: MURPHY

DATE/TIME OF TEST: 07/13/2016 15:31:49

Diagnostics Check	OK	15:25
Air Blank	0.000	15:26
Control Test	0.078	15:26
Air Blank	0.000	15:26
Subject Sample # 1	0.219	15:27
Breath Volume	2.363	15:27
Air Blank	0.000	15:27
Air Blank	0.000	15:29
Subject Sample # 2	0.228	15:30
Breath Volume	2.625	15:30
Air Blank	0.000	15:30
Control Test	0.076	15:31
Air Blank	0.000	15:31
Diagnostics Check	OK	15:31

Data Download Date/Time: 07/22/2016 13:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000759

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARNER,JOSHUA,C
BREATH TEST OPERATOR AGENCY: Jackson County SO
ARRESTING OFFICER: MITCHELL,WESLEY
ARRESTING OFFICER AGENCY: Jackson County SO
LAST AGENCY INSPECTION: 07/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A3
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 13:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:36

SUBJECT'S LAST NAME: KERN

DATE/TIME OF TEST: 07/23/2016 15:09:29

Diagnostics Check	OK	15:03	
Air Blank	0.000	15:04	
Control Test	0.076	15:04	
Air Blank	0.000	15:04	
Subject Sample	# 1 0.235	15:05	
Breath Volume	2.335	15:05	
Air Blank	0.000	15:06	
Air Blank	0.000	15:07	
Subject Sample	# 2	15:08	RFI Detect
Breath Volume	0.000	15:08	
Air Blank	0.000	15:08	
Control Test	0.076	15:08	
Air Blank	0.000	15:09	
Diagnostics Check	OK	15:09	

Data Download Date/Time: 08/14/2016 15:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000759

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARNER,JOSHUA,C
BREATH TEST OPERATOR AGENCY: Jackson County SO
ARRESTING OFFICER: MITCHELL,WESLEY
ARRESTING OFFICER AGENCY: Jackson County SO
LAST AGENCY INSPECTION: 07/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A3
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 13:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:36

SUBJECT'S LAST NAME: KERN

DATE/TIME OF TEST: 07/23/2016 15:18:39

Diagnostics Check	OK	15:12	
Air Blank	0.000	15:12	
Control Test	0.078	15:12	
Air Blank	0.000	15:13	
Subject Sample # 1	0.247	15:13	
Breath Volume	2.636	15:13	
Air Blank	0.000	15:14	
Air Blank	0.000	15:16	
Subject Sample # 2	0.227	15:16	
Breath Volume	3.484	15:16	
Air Blank	0.000	15:17	
Control Test	0.074	15:18	Control Outside Tolerance
Air Blank	0.000	15:18	

Data Download Date/Time: 08/14/2016 15:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000770

INSTRUMENT REGISTERED TO: HAMILTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,SONYA,C
BREATH TEST OPERATOR AGENCY: Jasper PD
ARRESTING OFFICER: HERNANDEZ,SONYA
ARRESTING OFFICER AGENCY: Jasper PD
LAST AGENCY INSPECTION: 06/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06415080A2
DRY GAS EXPIRY DATE: 04/05/2017
VIOLATION CODE: DUI
ARREST TIME: 19:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: LUCIANI

DATE/TIME OF TEST: 07/07/2016 20:29:25

Diagnostics Check	OK	20:26	
Air Blank	0.000	20:26	
Control Test	0.078	20:27	
Air Blank	0.000	20:27	
Subject Sample	# 1	20:28	Subject Test Refused
Breath Volume	0.000	20:28	
Air Blank	0.000	20:28	
Control Test	0.078	20:28	
Air Blank	0.000	20:29	
Diagnostics Check	OK	20:29	

Data Download Date/Time: 07/31/2016 20:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000770

INSTRUMENT REGISTERED TO: HAMILTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORGAN,NATHAN,A
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: MORGAN,NATHAN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 06/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06415080A2
DRY GAS EXPIRY DATE: 04/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:49

SUBJECT'S LAST NAME: RICO FIGUEROA

DATE/TIME OF TEST: 07/22/2016 22:24:50

Diagnostics Check	OK	22:09	
Air Blank	0.000	22:10	
Control Test	0.078	22:10	
Air Blank	0.000	22:11	
Subject Sample	# 1 0.191	22:12	
Breath Volume	1.656	22:12	
Air Blank	0.000	22:12	
Air Blank	0.000	22:14	
Subject Sample	# 2 0.162	22:17	Slope Not Level
Breath Volume	1.535	22:17	
Air Blank	0.000	22:18	
Air Blank	0.000	22:20	
Subject Sample	# 3	22:23	No Sample Provided
Breath Volume	0.000	22:23	
Air Blank	0.000	22:23	
Control Test	0.077	22:24	
Air Blank	0.000	22:24	
Diagnostics Check	OK	22:24	

Data Download Date/Time: 07/31/2016 20:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000770

INSTRUMENT REGISTERED TO: HAMILTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,SONYA,C
BREATH TEST OPERATOR AGENCY: Jasper PD
ARRESTING OFFICER: HERNANDEZ,SONYA
ARRESTING OFFICER AGENCY: Jasper PD
LAST AGENCY INSPECTION: 06/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06415080A2
DRY GAS EXPIRY DATE: 04/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: SANTIGOMEZ

DATE/TIME OF TEST: 07/24/2016 22:39:10

Diagnostics Check		OK	22:32
Air Blank		0.000	22:32
Control Test		0.078	22:33
Air Blank		0.000	22:33
Subject Sample	# 1	0.250	22:34
Breath Volume		2.015	22:34
Air Blank		0.000	22:35
Air Blank		0.000	22:37
Subject Sample	# 2	0.240	22:37
Breath Volume		1.781	22:37
Air Blank		0.000	22:38
Control Test		0.078	22:38
Air Blank		0.000	22:39
Diagnostics Check		OK	22:39

Data Download Date/Time: 07/31/2016 20:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000771

INSTRUMENT REGISTERED TO: PERRY POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CANNON,CHAD,M
BREATH TEST OPERATOR AGENCY: Perry PD
ARRESTING OFFICER: CANNON,CHAD
ARRESTING OFFICER AGENCY: Perry PD
LAST AGENCY INSPECTION: 06/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG425202
DRY GAS EXPIRY DATE: 09/09/2016
VIOLATION CODE: DUI
ARREST TIME: 21:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:21

SUBJECT'S LAST NAME: MATHIS

DATE/TIME OF TEST: 07/06/2016 21:52:44

Diagnostics Check	OK	21:45
Air Blank	0.000	21:46
Control Test	0.078	21:46
Air Blank	0.000	21:47
Subject Sample # 1	0.174	21:48
Breath Volume	2.136	21:48
Air Blank	0.000	21:49
Air Blank	0.000	21:50
Subject Sample # 2	0.156	21:51
Breath Volume	1.140	21:51
Air Blank	0.000	21:51
Control Test	0.075	21:52
Air Blank	0.000	21:52
Diagnostics Check	OK	21:52

Data Download Date/Time: 07/27/2016 16:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIGHTOWER,CONNIE,A
BREATH TEST OPERATOR AGENCY: Lake City PD
ARRESTING OFFICER: ROBINSON,STUART
ARRESTING OFFICER AGENCY: Lake City PD
LAST AGENCY INSPECTION: 06/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: BLACKWELL

DATE/TIME OF TEST: 07/02/2016 23:13:52

Diagnostics Check	OK	23:07
Air Blank	0.000	23:07
Control Test	0.082	23:08
Air Blank	0.000	23:08
Subject Sample # 1	0.148	23:09
Breath Volume	1.816	23:09
Air Blank	0.000	23:09
Air Blank	0.000	23:11
Subject Sample # 2	0.146	23:12
Breath Volume	1.367	23:12
Air Blank	0.000	23:12
Control Test	0.081	23:13
Air Blank	0.000	23:13
Diagnostics Check	OK	23:13

Data Download Date/Time: 07/19/2016 17:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OGLESBY, MICHAEL, L
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: OGLESBY, MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 06/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: BRADSHAW

DATE/TIME OF TEST: 07/04/2016 23:45:31

Diagnostics Check	OK	23:30	
Air Blank	0.000	23:30	
Control Test	0.082	23:31	
Air Blank	0.000	23:31	
Subject Sample	# 1 0.175	23:32	
Breath Volume	1.523	23:32	
Air Blank	0.000	23:33	
Air Blank	0.000	23:35	
Subject Sample	# 2	23:38	No Sample Provided
Breath Volume	0.000	23:38	
Air Blank	0.000	23:38	
Air Blank	0.000	23:41	
Subject Sample	# 3	23:44	No Sample Provided
Breath Volume	0.000	23:44	
Air Blank	0.000	23:44	
Control Test	0.082	23:44	
Air Blank	0.000	23:45	
Diagnostics Check	OK	23:45	

Data Download Date/Time: 07/19/2016 17:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIGHTOWER,CONNIE,A
BREATH TEST OPERATOR AGENCY: Lake City PD
ARRESTING OFFICER: MICHAELS,P
ARRESTING OFFICER AGENCY: Lake City PD
LAST AGENCY INSPECTION: 06/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 13:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:58

SUBJECT'S LAST NAME: CHROME

DATE/TIME OF TEST: 07/11/2016 14:28:14

Diagnostics Check	OK	14:21
Air Blank	0.000	14:22
Control Test	0.082	14:22
Air Blank	0.000	14:23
Subject Sample # 1	0.163	14:23
Breath Volume	2.718	14:23
Air Blank	0.000	14:24
Air Blank	0.000	14:26
Subject Sample # 2	0.168	14:26
Breath Volume	2.367	14:26
Air Blank	0.000	14:27
Control Test	0.079	14:27
Air Blank	0.000	14:28
Diagnostics Check	OK	14:28

Data Download Date/Time: 07/19/2016 17:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORGAN,NATHAN,A
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: MORGAN,NATHAN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 06/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 18:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:04

SUBJECT'S LAST NAME: GRCIAN

DATE/TIME OF TEST: 07/13/2016 19:30:59

Diagnostics Check	OK	19:24
Air Blank	0.000	19:25
Control Test	0.081	19:25
Air Blank	0.000	19:25
Subject Sample # 1	0.000	19:26
Breath Volume	2.542	19:26
Air Blank	0.000	19:27
Air Blank	0.000	19:29
Subject Sample # 2	0.000	19:29
Breath Volume	1.906	19:29
Air Blank	0.000	19:30
Control Test	0.081	19:30
Air Blank	0.000	19:30
Diagnostics Check	OK	19:30

Data Download Date/Time: 07/19/2016 17:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPENCE,KEITH,D
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: SPENCE,KEITH
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 06/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: REGISTER

DATE/TIME OF TEST: 07/16/2016 03:37:57

Diagnostics Check	OK	03:28
Air Blank	0.000	03:29
Control Test	0.082	03:29
Air Blank	0.000	03:30
Subject Sample # 1	0.089	03:33
Breath Volume	2.492	03:33
Air Blank	0.000	03:33
Air Blank	0.000	03:35
Subject Sample # 2	0.084	03:36
Breath Volume	1.480	03:36
Air Blank	0.000	03:37
Control Test	0.081	03:37
Air Blank	0.000	03:37
Diagnostics Check	OK	03:37

Data Download Date/Time: 07/19/2016 17:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,ALEXIS,
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: BIRCHARD,MARK
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 07/19/2016

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 17:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:28

SUBJECT'S LAST NAME: BRANHAM

DATE/TIME OF TEST: 07/19/2016 18:59:01

Diagnostics Check	OK	18:52
Air Blank	0.000	18:52
Control Test	0.081	18:53
Air Blank	0.000	18:53
Subject Sample # 1	0.138	18:54
Breath Volume	3.335	18:54
Air Blank	0.000	18:54
Air Blank	0.000	18:56
Subject Sample # 2	0.143	18:57
Breath Volume	2.503	18:57
Air Blank	0.000	18:58
Control Test	0.080	18:58
Air Blank	0.000	18:58
Diagnostics Check	OK	18:59

Data Download Date/Time: 08/18/2016 20:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,ALEXIS,
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: HERNANDEZ,ALEXIS
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 07/19/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:36

SUBJECT'S LAST NAME: BENZ

DATE/TIME OF TEST: 07/23/2016 23:04:09

Diagnostics Check	OK	22:57
Air Blank	0.000	22:57
Control Test	0.082	22:57
Air Blank	0.000	22:58
Subject Sample # 1	0.167	22:59
Breath Volume	2.257	22:59
Air Blank	0.000	22:59
Air Blank	0.000	23:01
Subject Sample # 2	0.158	23:02
Breath Volume	2.363	23:02
Air Blank	0.000	23:03
Control Test	0.079	23:03
Air Blank	0.000	23:04
Diagnostics Check	OK	23:04

Data Download Date/Time: 08/18/2016 20:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLE,DANDRA,M
BREATH TEST OPERATOR AGENCY: Columbia County SO
ARRESTING OFFICER: COLE,DANDRA
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 07/19/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: ALVEY

DATE/TIME OF TEST: 07/24/2016 02:56:27

Diagnostics Check	OK	02:49
Air Blank	0.000	02:50
Control Test	0.082	02:50
Air Blank	0.000	02:51
Subject Sample # 1	0.174	02:51
Breath Volume	3.546	02:51
Air Blank	0.000	02:52
Air Blank	0.000	02:54
Subject Sample # 2	0.155	02:54
Breath Volume	1.621	02:54
Air Blank	0.000	02:55
Control Test	0.080	02:55
Air Blank	0.000	02:56
Diagnostics Check	OK	02:56

Data Download Date/Time: 08/18/2016 20:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIGHTOWER,CONNIE,A
BREATH TEST OPERATOR AGENCY: Lake City PD
ARRESTING OFFICER: HIGHTOWER,CONNIE
ARRESTING OFFICER AGENCY: Lake City PD
LAST AGENCY INSPECTION: 07/19/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:56

SUBJECT'S LAST NAME: VINCENT

DATE/TIME OF TEST: 07/24/2016 03:19:37

Diagnostics Check	OK	03:16	
Air Blank	0.000	03:17	
Control Test	0.082	03:17	
Air Blank	0.000	03:18	
Subject Sample	# 1	03:18	Subject Test Refused
Breath Volume	0.000	03:18	
Air Blank	0.000	03:18	
Control Test	0.081	03:19	
Air Blank	0.000	03:19	
Diagnostics Check	OK	03:19	

Data Download Date/Time: 08/18/2016 20:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLE,DANDRA,M
BREATH TEST OPERATOR AGENCY: Columbia County SO
ARRESTING OFFICER: GUERRY,CHAD
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 07/19/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:46

SUBJECT'S LAST NAME: LAPORTE

DATE/TIME OF TEST: 07/31/2016 03:17:18

Diagnostics Check		OK	03:07
Air Blank		0.000	03:07
Control Test		0.082	03:07
Air Blank		0.000	03:08
Subject Sample	# 1	0.143	03:09
Breath Volume		1.855	03:09
Air Blank		0.000	03:10
Air Blank		0.000	03:12
Subject Sample	# 2	0.173	03:12
Breath Volume		2.324	03:12
Air Blank		0.000	03:13
Air Blank		0.000	03:15
Subject Sample	# 3	0.168	03:15
Breath Volume		1.898	03:15
Air Blank		0.000	03:16
Control Test		0.076	03:16
Air Blank		0.000	03:17
Diagnostics Check		OK	03:17

Data Download Date/Time: 08/18/2016 20:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000776

INSTRUMENT REGISTERED TO: LAFAYETTE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEAS,ALFRED,A
BREATH TEST OPERATOR AGENCY: Lafayette County SO
ARRESTING OFFICER: DEAS,ALFRED
ARRESTING OFFICER AGENCY: Lafayette County SO
LAST AGENCY INSPECTION: 06/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 17:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:40

SUBJECT'S LAST NAME: REID

DATE/TIME OF TEST: 07/02/2016 19:16:17

Diagnostics Check	OK	19:12	
Air Blank	0.000	19:13	
Control Test	0.079	19:13	
Air Blank	0.000	19:14	
Subject Sample	# 1	19:14	RFI Detect
Breath Volume	0.675	19:14	
Air Blank	0.000	19:15	
Control Test	0.079	19:15	
Air Blank	0.000	19:16	
Diagnostics Check	OK	19:16	

Data Download Date/Time: 08/10/2016 10:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000776

INSTRUMENT REGISTERED TO: LAFAYETTE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEAS,ALFRED,A
BREATH TEST OPERATOR AGENCY: Lafayette County SO
ARRESTING OFFICER: DEAS,ALFRED
ARRESTING OFFICER AGENCY: Lafayette County SO
LAST AGENCY INSPECTION: 06/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 17:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:40

SUBJECT'S LAST NAME: REID

DATE/TIME OF TEST: 07/02/2016 19:25:28

Diagnostics Check	OK	19:19
Air Blank	0.000	19:20
Control Test	0.079	19:20
Air Blank	0.000	19:20
Subject Sample # 1	0.073	19:21
Breath Volume	1.210	19:21
Air Blank	0.000	19:21
Air Blank	0.000	19:23
Subject Sample # 2	0.069	19:24
Breath Volume	1.203	19:24
Air Blank	0.000	19:24
Control Test	0.079	19:24
Air Blank	0.000	19:25
Diagnostics Check	OK	19:25

Data Download Date/Time: 08/10/2016 10:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000779

INSTRUMENT REGISTERED TO: MADISON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,JAMES,D
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: SMITH,JAMES
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 07/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: STOUT

DATE/TIME OF TEST: 07/16/2016 01:36:10

Diagnostics Check	OK	01:33	
Air Blank	0.000	01:33	
Control Test	0.081	01:34	
Air Blank	0.000	01:34	
Subject Sample	# 1	01:34	Subject Test Refused
Breath Volume	0.000	01:34	
Air Blank	0.000	01:35	
Control Test	0.081	01:35	
Air Blank	0.000	01:36	
Diagnostics Check	OK	01:36	

Data Download Date/Time: 08/17/2016 11:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000779

INSTRUMENT REGISTERED TO: MADISON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEWEY,JARED,K
BREATH TEST OPERATOR AGENCY: Madison County SO
ARRESTING OFFICER: DEWEY,JARED
ARRESTING OFFICER AGENCY: Madison County SO
LAST AGENCY INSPECTION: 07/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 18:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:55

SUBJECT'S LAST NAME: GORDON

DATE/TIME OF TEST: 07/16/2016 19:22:02

Diagnostics Check	OK	19:15
Air Blank	0.000	19:16
Control Test	0.081	19:16
Air Blank	0.000	19:17
Subject Sample # 1	0.131	19:17
Breath Volume	3.785	19:17
Air Blank	0.000	19:18
Air Blank	0.000	19:20
Subject Sample # 2	0.130	19:20
Breath Volume	3.640	19:20
Air Blank	0.000	19:21
Control Test	0.079	19:21
Air Blank	0.000	19:21
Diagnostics Check	OK	19:22

Data Download Date/Time: 08/17/2016 11:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000779

INSTRUMENT REGISTERED TO: MADISON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLEMENT,JOSEPH,E
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: CLEMENT,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 07/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:56

SUBJECT'S LAST NAME: VALDEZ

DATE/TIME OF TEST: 07/30/2016 01:31:20

Diagnostics Check	OK	01:24
Air Blank	0.000	01:25
Control Test	0.079	01:25
Air Blank	0.000	01:26
Subject Sample # 1	0.244	01:26
Breath Volume	3.367	01:26
Air Blank	0.000	01:27
Air Blank	0.000	01:29
Subject Sample # 2	0.245	01:29
Breath Volume	4.125	01:29
Air Blank	0.000	01:30
Control Test	0.079	01:30
Air Blank	0.000	01:31
Diagnostics Check	OK	01:31

Data Download Date/Time: 08/17/2016 11:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000779

INSTRUMENT REGISTERED TO: MADISON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,JAMES,D
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: SMITH,JAMES
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 07/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:14

SUBJECT'S LAST NAME: DIXON

DATE/TIME OF TEST: 07/30/2016 20:44:53

Diagnostics Check	OK	20:35
Air Blank	0.000	20:36
Control Test	0.081	20:36
Air Blank	0.000	20:37
Subject Sample # 1	0.162	20:38
Breath Volume	2.187	20:38
Air Blank	0.000	20:39
Air Blank	0.000	20:41
Subject Sample # 2	0.156	20:43
Breath Volume	2.214	20:43
Air Blank	0.000	20:43
Control Test	0.076	20:44
Air Blank	0.000	20:44
Diagnostics Check	OK	20:44

Data Download Date/Time: 08/17/2016 11:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000780

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEE, WILLIAM, C
BREATH TEST OPERATOR AGENCY: Suwannee County SO
ARRESTING OFFICER: HINGSON, THOMAS
ARRESTING OFFICER AGENCY: Live Oak PD
LAST AGENCY INSPECTION: 06/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06616080A2
DRY GAS EXPIRY DATE: 04/05/2018
VIOLATION CODE: DUI
ARREST TIME: 13:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:30

SUBJECT'S LAST NAME: VALQUEZ

DATE/TIME OF TEST: 07/01/2016 14:21:23

Diagnostics Check	OK	14:14
Air Blank	0.000	14:15
Control Test	0.082	14:15
Air Blank	0.000	14:16
Subject Sample # 1	0.236	14:16
Breath Volume	3.003	14:16
Air Blank	0.000	14:17
Air Blank	0.000	14:19
Subject Sample # 2	0.233	14:19
Breath Volume	2.480	14:19
Air Blank	0.000	14:20
Control Test	0.081	14:20
Air Blank	0.000	14:21
Diagnostics Check	OK	14:21

Data Download Date/Time: 07/31/2016 03:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000780

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY, MICHAEL, S
BREATH TEST OPERATOR AGENCY: Suwannee County SO
ARRESTING OFFICER: WATSON, CLAYTON
ARRESTING OFFICER AGENCY: Suwannee County SO
LAST AGENCY INSPECTION: 06/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06616080A2
DRY GAS EXPIRY DATE: 04/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:08

SUBJECT'S LAST NAME: SPINKS

DATE/TIME OF TEST: 07/02/2016 04:38:27

Diagnostics Check	OK	04:29
Air Blank	0.000	04:30
Control Test	0.082	04:30
Air Blank	0.000	04:31
Subject Sample # 1	0.184	04:33
Breath Volume	2.875	04:33
Air Blank	0.000	04:34
Air Blank	0.000	04:36
Subject Sample # 2	0.177	04:36
Breath Volume	2.781	04:36
Air Blank	0.000	04:37
Control Test	0.081	04:37
Air Blank	0.000	04:38
Diagnostics Check	OK	04:38

Data Download Date/Time: 07/31/2016 03:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000786

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DICIENZO,FRANCA,R
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: MORALES,DAVID
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf
LAST AGENCY INSPECTION: 06/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 01:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: VALENZUELA

DATE/TIME OF TEST: 07/09/2016 01:50:31

Diagnostics Check	OK	01:44
Air Blank	0.000	01:44
Control Test	0.078	01:45
Air Blank	0.000	01:45
Subject Sample # 1	0.132	01:46
Breath Volume	2.078	01:46
Air Blank	0.000	01:46
Air Blank	0.000	01:48
Subject Sample # 2	0.126	01:49
Breath Volume	1.945	01:49
Air Blank	0.000	01:49
Control Test	0.079	01:49
Air Blank	0.000	01:50
Diagnostics Check	OK	01:50

Data Download Date/Time: 07/26/2016 11:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TATRO,KERSTAN,J
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: KIDD,JERMEL
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 07/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:52

SUBJECT'S LAST NAME: REYNOLDS

DATE/TIME OF TEST: 07/14/2016 00:15:18

Diagnostics Check	OK	00:12	
Air Blank	0.000	00:12	
Control Test	0.077	00:13	
Air Blank	0.000	00:13	
Subject Sample	# 1	00:13	Subject Test Refused
Breath Volume	0.000	00:13	
Air Blank	0.000	00:14	
Control Test	0.077	00:14	
Air Blank	0.000	00:15	
Diagnostics Check	OK	00:15	

Data Download Date/Time: 08/24/2016 20:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOTWALA, PRASHANT,
BREATH TEST OPERATOR AGENCY: OTHER Fed
ARRESTING OFFICER: LOTWALA, PRASHANT
ARRESTING OFFICER AGENCY: OTHER Fed
LAST AGENCY INSPECTION: 07/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 16:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:08

SUBJECT'S LAST NAME: SHINGLER

DATE/TIME OF TEST: 07/16/2016 20:36:35

Diagnostics Check	OK	20:30
Air Blank	0.000	20:30
Control Test	0.078	20:30
Air Blank	0.000	20:31
Subject Sample # 1	0.181	20:32
Breath Volume	1.378	20:32
Air Blank	0.000	20:32
Air Blank	0.000	20:34
Subject Sample # 2	0.181	20:35
Breath Volume	1.726	20:35
Air Blank	0.000	20:35
Control Test	0.078	20:36
Air Blank	0.000	20:36
Diagnostics Check	OK	20:36

Data Download Date/Time: 08/24/2016 20:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERSON,THOMAS,C
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: ROBERSON,THOMAS
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 07/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:17

SUBJECT'S LAST NAME: FOURROUX

DATE/TIME OF TEST: 07/16/2016 23:43:52

Diagnostics Check	OK	23:41	
Air Blank	0.000	23:41	
Control Test	0.077	23:41	
Air Blank	0.000	23:42	
Subject Sample	# 1	23:42	Subject Test Refused
Breath Volume	0.000	23:42	
Air Blank	0.000	23:42	
Control Test	0.076	23:43	
Air Blank	0.000	23:43	
Diagnostics Check	OK	23:43	

Data Download Date/Time: 08/24/2016 20:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARD,BENJAMIN,D
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: WARD,BENJAMIN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 07/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 15:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:37

SUBJECT'S LAST NAME: GRIESHOP

DATE/TIME OF TEST: 07/17/2016 17:09:38

Diagnostics Check	OK	17:03
Air Blank	0.000	17:03
Control Test	0.077	17:04
Air Blank	0.000	17:04
Subject Sample # 1	0.287	17:05
Breath Volume	1.640	17:05
Air Blank	0.000	17:05
Air Blank	0.000	17:07
Subject Sample # 2	0.285	17:08
Breath Volume	2.281	17:08
Air Blank	0.000	17:08
Control Test	0.076	17:09
Air Blank	0.000	17:09
Diagnostics Check	OK	17:09

Data Download Date/Time: 08/24/2016 20:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCAPIN,BRANDON,L
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: SCAPIN,BRANDON
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 07/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: KAESER

DATE/TIME OF TEST: 07/20/2016 01:33:17

Diagnostics Check	OK	01:30	
Air Blank	0.000	01:30	
Control Test	0.078	01:31	
Air Blank	0.000	01:31	
Subject Sample	# 1	01:31	Subject Test Refused
Breath Volume	0.000	01:31	
Air Blank	0.000	01:32	
Control Test	0.078	01:32	
Air Blank	0.000	01:33	
Diagnostics Check	OK	01:33	

Data Download Date/Time: 08/24/2016 20:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NGUYEN,TAI,T
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: NGUYEN,TAI
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 07/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 11:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:45

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 07/23/2016 12:23:08

Diagnostics Check	OK	12:15	
Air Blank	0.000	12:15	
Control Test	0.076	12:16	
Air Blank	0.000	12:16	
Subject Sample	# 1 0.195	12:17	
Breath Volume	1.160	12:17	
Air Blank	0.000	12:17	
Air Blank	0.000	12:19	
Subject Sample	# 2 0.194	12:20	
Breath Volume	1.652	12:20	
Air Blank	0.000	12:21	
Control Test	0.077	12:22	
Air Blank	0.000	12:22	
Diagnostics Check	Fail	12:22	Temperature Regulation Test
Air Blank	0.000	12:23	

Data Download Date/Time: 08/24/2016 20:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NGUYEN,TAI,T
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: NGUYEN,TAI
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 07/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 11:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:50

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 07/23/2016 12:33:45

Diagnostics Check	Fail	12:33	Temperature Regulation Test
Air Blank	0.000	12:33	

Data Download Date/Time: 08/24/2016 20:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NGUYEN,TAI,T
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: NGUYEN,TAI
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 07/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 11:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:50

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 07/23/2016 12:36:34

Diagnostics Check	Fail	12:36	Temperature Regulation Test
Air Blank	0.000	12:36	

Data Download Date/Time: 08/24/2016 20:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NGUYEN,TAI,T
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: NGUYEN,TAI
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 07/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 11:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:50

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 07/23/2016 12:39:36

Diagnostics Check	Fail	12:39	Temperature Regulation Test
Air Blank	0.000	12:39	

Data Download Date/Time: 08/24/2016 20:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NGUYEN,TAI,T
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: NGUYEN,TAI
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 07/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 11:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:55

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 07/23/2016 14:26:57

Diagnostics Check	OK	14:19
Air Blank	0.000	14:20
Control Test	0.077	14:20
Air Blank	0.000	14:21
Subject Sample # 1	0.154	14:22
Breath Volume	1.285	14:22
Air Blank	0.000	14:23
Air Blank	0.000	14:24
Subject Sample # 2	0.141	14:25
Breath Volume	1.808	14:25
Air Blank	0.000	14:26
Control Test	0.075	14:26
Air Blank	0.000	14:26
Diagnostics Check	OK	14:26

Data Download Date/Time: 08/24/2016 20:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TATRO,KERSTAN,J
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: TATRO,KERSTAN
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 07/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: OREA

DATE/TIME OF TEST: 07/24/2016 01:44:39

Diagnostics Check	OK	01:41	
Air Blank	0.000	01:42	
Control Test	0.078	01:42	
Air Blank	0.000	01:43	
Subject Sample	# 1	01:43	Subject Test Refused
Breath Volume	0.000	01:43	
Air Blank	0.000	01:43	
Control Test	0.078	01:44	
Air Blank	0.000	01:44	
Diagnostics Check	OK	01:44	

Data Download Date/Time: 08/24/2016 20:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCAPIN,BRANDON,L
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: SCAPIN,BRANDON
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 07/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: MEAD

DATE/TIME OF TEST: 07/24/2016 05:00:17

Diagnostics Check	OK	04:57	
Air Blank	0.000	04:57	
Control Test	0.077	04:58	
Air Blank	0.000	04:58	
Subject Sample	# 1	04:58	Subject Test Refused
Breath Volume	0.000	04:58	
Air Blank	0.000	04:59	
Control Test	0.078	04:59	
Air Blank	0.000	05:00	
Diagnostics Check	OK	05:00	

Data Download Date/Time: 08/24/2016 20:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCAPIN, BRANDON, L
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: CUSHING, MATTHEW
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 07/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: BUI
ARREST TIME: 01:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:27

SUBJECT'S LAST NAME: FRANKLIN

DATE/TIME OF TEST: 07/25/2016 02:53:45

Diagnostics Check	OK	02:47
Air Blank	0.000	02:47
Control Test	0.078	02:47
Air Blank	0.000	02:48
Subject Sample # 1	0.181	02:49
Breath Volume	2.332	02:49
Air Blank	0.000	02:49
Air Blank	0.000	02:51
Subject Sample # 2	0.175	02:52
Breath Volume	1.894	02:52
Air Blank	0.000	02:52
Control Test	0.077	02:53
Air Blank	0.000	02:53
Diagnostics Check	OK	02:53

Data Download Date/Time: 08/24/2016 20:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000790

INSTRUMENT REGISTERED TO: VERO BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOHMAN,JENNIFER,L
BREATH TEST OPERATOR AGENCY: Vero Beach PD
ARRESTING OFFICER: LOHMAN,JENNIFER
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 06/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 16014080A2
DRY GAS EXPIRY DATE: 07/05/2016
VIOLATION CODE: DUI
ARREST TIME: 18:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:23

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 07/01/2016 18:50:01

Diagnostics Check	OK	18:44
Air Blank	0.000	18:44
Control Test	0.079	18:44
Air Blank	0.000	18:45
Subject Sample # 1	0.000	18:45
Breath Volume	2.066	18:45
Air Blank	0.000	18:46
Air Blank	0.000	18:48
Subject Sample # 2	0.000	18:48
Breath Volume	2.128	18:48
Air Blank	0.000	18:49
Control Test	0.079	18:49
Air Blank	0.000	18:49
Diagnostics Check	OK	18:50

Data Download Date/Time: 07/04/2016 06:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000790

INSTRUMENT REGISTERED TO: VERO BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOMBARDO,RICHARD,V
BREATH TEST OPERATOR AGENCY: Vero Beach PD
ARRESTING OFFICER: VASQUEZ,CHRIS
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 07/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 16014080A2
DRY GAS EXPIRY DATE: 07/05/2016
VIOLATION CODE: DUI
ARREST TIME: 01:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:06

SUBJECT'S LAST NAME: WILLER

DATE/TIME OF TEST: 07/10/2016 01:39:26

Diagnostics Check	OK	01:32
Air Blank	0.000	01:33
Control Test	0.079	01:33
Air Blank	0.000	01:34
Subject Sample # 1	0.100	01:34
Breath Volume	3.128	01:34
Air Blank	0.000	01:35
Air Blank	0.000	01:37
Subject Sample # 2	0.095	01:37
Breath Volume	4.117	01:37
Air Blank	0.000	01:38
Control Test	0.078	01:38
Air Blank	0.000	01:39
Diagnostics Check	OK	01:39

Data Download Date/Time: 08/29/2016 02:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000790

INSTRUMENT REGISTERED TO: VERO BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOMBARDO,RICHARD,V
BREATH TEST OPERATOR AGENCY: Vero Beach PD
ARRESTING OFFICER: VASQUEZ,CHRIS
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 07/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 16014080A2
DRY GAS EXPIRY DATE: 07/05/2016
VIOLATION CODE: DUI
ARREST TIME: 02:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:54

SUBJECT'S LAST NAME: COSTANZO

DATE/TIME OF TEST: 07/11/2016 03:44:05

Diagnostics Check	OK	03:37
Air Blank	0.000	03:37
Control Test	0.079	03:38
Air Blank	0.000	03:38
Subject Sample # 1	0.324	03:39
Breath Volume	5.050	03:39
Air Blank	0.000	03:40
Air Blank	0.000	03:41
Subject Sample # 2	0.321	03:42
Breath Volume	4.835	03:42
Air Blank	0.000	03:43
Control Test	0.077	03:43
Air Blank	0.000	03:44
Diagnostics Check	OK	03:44

Data Download Date/Time: 08/29/2016 02:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASSERLY,STEPHEN,R
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: DIONISIO,DAVID
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 06/17/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35213080A3
DRY GAS EXPIRY DATE: 01/01/2016
VIOLATION CODE: DUI
ARREST TIME: 01:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:08

SUBJECT'S LAST NAME: MEEK

DATE/TIME OF TEST: 07/02/2016 02:36:35

Diagnostics Check	OK	02:29
Air Blank	0.000	02:30
Control Test	0.079	02:30
Air Blank	0.000	02:31
Subject Sample # 1	0.229	02:31
Breath Volume	1.925	02:31
Air Blank	0.000	02:32
Air Blank	0.000	02:34
Subject Sample # 2	0.219	02:34
Breath Volume	1.363	02:34
Air Blank	0.000	02:35
Control Test	0.077	02:35
Air Blank	0.000	02:36
Diagnostics Check	OK	02:36

Data Download Date/Time: 07/12/2016 12:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEWIS,JEREMY,A
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: WYATT,PJ
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 06/17/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35213080A3
DRY GAS EXPIRY DATE: 01/01/2016
VIOLATION CODE: DUI
ARREST TIME: 04:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:20

SUBJECT'S LAST NAME: CRUZ ESCOBAR

DATE/TIME OF TEST: 07/02/2016 05:46:14

Diagnostics Check		OK	05:40
Air Blank		0.000	05:40
Control Test		0.079	05:40
Air Blank		0.000	05:41
Subject Sample	# 1	0.107	05:41
Breath Volume		1.929	05:41
Air Blank		0.000	05:42
Air Blank		0.000	05:44
Subject Sample	# 2	0.109	05:44
Breath Volume		2.171	05:44
Air Blank		0.000	05:45
Control Test		0.079	05:45
Air Blank		0.000	05:46
Diagnostics Check		OK	05:46

Data Download Date/Time: 07/12/2016 12:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PATTERSON,KYLE,L
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: PATTERSON,KYLE
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 06/17/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35213080A3
DRY GAS EXPIRY DATE: 01/01/2016
VIOLATION CODE: BUI
ARREST TIME: 18:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:30

SUBJECT'S LAST NAME: COBERLY

DATE/TIME OF TEST: 07/02/2016 18:56:42

Diagnostics Check	OK	18:50
Air Blank	0.000	18:51
Control Test	0.079	18:51
Air Blank	0.000	18:51
Subject Sample # 1	0.048	18:52
Breath Volume	1.363	18:52
Air Blank	0.000	18:52
Air Blank	0.000	18:54
Subject Sample # 2	0.048	18:55
Breath Volume	3.234	18:55
Air Blank	0.000	18:55
Control Test	0.078	18:56
Air Blank	0.000	18:56
Diagnostics Check	OK	18:56

Data Download Date/Time: 07/12/2016 12:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOGAN,TODD,A
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: HOGAN,TODD
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 06/17/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35213080A3
DRY GAS EXPIRY DATE: 01/01/2016
VIOLATION CODE: DUI
ARREST TIME: 21:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:41

SUBJECT'S LAST NAME: OBRIEN

DATE/TIME OF TEST: 07/02/2016 22:10:52

Diagnostics Check	OK	22:04
Air Blank	0.000	22:04
Control Test	0.079	22:04
Air Blank	0.000	22:05
Subject Sample # 1	0.119	22:06
Breath Volume	1.488	22:06
Air Blank	0.000	22:06
Air Blank	0.000	22:08
Subject Sample # 2	0.118	22:09
Breath Volume	1.367	22:09
Air Blank	0.000	22:09
Control Test	0.078	22:10
Air Blank	0.000	22:10
Diagnostics Check	OK	22:10

Data Download Date/Time: 07/12/2016 12:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANTON,NICHOLAS,A
BREATH TEST OPERATOR AGENCY: Fort Pierce PD
ARRESTING OFFICER: ANTON,NICHOLAS
ARRESTING OFFICER AGENCY: Fort Pierce PD
LAST AGENCY INSPECTION: 06/17/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35213080A3
DRY GAS EXPIRY DATE: 10/05/2016
VIOLATION CODE: DUI
ARREST TIME: 22:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: BUZZIZ

DATE/TIME OF TEST: 07/03/2016 00:01:09

Diagnostics Check	OK	23:54
Air Blank	0.000	23:55
Control Test	0.079	23:55
Air Blank	0.000	23:55
Subject Sample # 1	0.206	23:56
Breath Volume	1.500	23:56
Air Blank	0.000	23:57
Air Blank	0.000	23:58
Subject Sample # 2	0.195	23:59
Breath Volume	1.945	23:59
Air Blank	0.000	00:00
Control Test	0.079	00:00
Air Blank	0.000	00:01
Diagnostics Check	OK	00:01

Data Download Date/Time: 07/12/2016 12:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANTON,NICHOLAS,A
BREATH TEST OPERATOR AGENCY: Fort Pierce PD
ARRESTING OFFICER: ANTON,NICHOLAS
ARRESTING OFFICER AGENCY: Fort Pierce PD
LAST AGENCY INSPECTION: 06/17/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35213080A3
DRY GAS EXPIRY DATE: 10/05/2016
VIOLATION CODE: DUI
ARREST TIME: 03:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: VASQUEZ

DATE/TIME OF TEST: 07/03/2016 04:52:56

Diagnostics Check	OK	04:45
Air Blank	0.000	04:45
Control Test	0.080	04:46
Air Blank	0.000	04:46
Subject Sample # 1	0.166	04:47
Breath Volume	1.480	04:47
Air Blank	0.000	04:48
Air Blank	0.000	04:50
Subject Sample # 2	0.153	04:51
Breath Volume	1.625	04:51
Air Blank	0.000	04:52
Control Test	0.079	04:52
Air Blank	0.000	04:52
Diagnostics Check	OK	04:52

Data Download Date/Time: 07/12/2016 12:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PATTERSON,KYLE,L
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: SHERMETARO,MARTIN
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 06/17/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35213080A3
DRY GAS EXPIRY DATE: 01/01/2016
VIOLATION CODE: BUI
ARREST TIME: 17:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:00

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 07/03/2016 18:27:40

Diagnostics Check	OK	18:21
Air Blank	0.000	18:21
Control Test	0.078	18:22
Air Blank	0.000	18:22
Subject Sample # 1	0.137	18:23
Breath Volume	2.183	18:23
Air Blank	0.000	18:23
Air Blank	0.000	18:25
Subject Sample # 2	0.128	18:25
Breath Volume	1.722	18:25
Air Blank	0.000	18:26
Control Test	0.081	18:27
Air Blank	0.000	18:27
Diagnostics Check	OK	18:27

Data Download Date/Time: 07/12/2016 12:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PATTERSON,KYLE,L
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: PAYNE,JAMES
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 06/17/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35213080A3
DRY GAS EXPIRY DATE: 01/01/2016
VIOLATION CODE: BUI
ARREST TIME: 20:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:40

SUBJECT'S LAST NAME: HARTMAN

DATE/TIME OF TEST: 07/03/2016 21:25:34

Diagnostics Check	OK	21:19
Air Blank	0.000	21:19
Control Test	0.078	21:19
Air Blank	0.000	21:20
Subject Sample # 1	0.187	21:20
Breath Volume	1.734	21:20
Air Blank	0.000	21:21
Air Blank	0.000	21:23
Subject Sample # 2	0.193	21:23
Breath Volume	2.332	21:23
Air Blank	0.000	21:24
Control Test	0.075	21:24
Air Blank	0.000	21:25
Diagnostics Check	OK	21:25

Data Download Date/Time: 07/12/2016 12:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DRAWDY,LAINE,A
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: DRAWDY,LAINE
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 06/17/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35213080A3
DRY GAS EXPIRY DATE: 01/01/2016
VIOLATION CODE: DUI
ARREST TIME: 22:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:15

SUBJECT'S LAST NAME: GUTIERREZ ZUNIGA

DATE/TIME OF TEST: 07/03/2016 22:49:05

Diagnostics Check	OK	22:39	
Air Blank	0.000	22:39	
Control Test	0.079	22:39	
Air Blank	0.000	22:40	
Subject Sample	# 1	22:43	No Sample Provided
Breath Volume	0.000	22:43	
Air Blank	0.000	22:44	
Air Blank	0.000	22:45	
Subject Sample	# 2	22:47	Subject Test Refused
Breath Volume	0.000	22:47	
Air Blank	0.000	22:48	
Control Test	0.077	22:48	
Air Blank	0.000	22:49	
Diagnostics Check	OK	22:49	

Data Download Date/Time: 07/12/2016 12:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DRAWDY,LAINE,A
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: DRAWDY,LAINE
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 06/17/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35213080A3
DRY GAS EXPIRY DATE: 01/01/2016
VIOLATION CODE: DUI
ARREST TIME: 01:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: MIZRACH

DATE/TIME OF TEST: 07/04/2016 03:02:36

Diagnostics Check	OK	02:49	
Air Blank	0.000	02:49	
Control Test	0.079	02:49	
Air Blank	0.000	02:50	
Subject Sample # 1	0.000	02:52	
Breath Volume	1.492	02:52	
Air Blank	0.000	02:52	
Air Blank	0.000	02:54	
Subject Sample # 2	0.000	02:57	Volume Not Met
Breath Volume	0.777	02:57	
Air Blank	0.000	02:58	
Air Blank	0.000	03:00	
Subject Sample # 3	0.000	03:01	
Breath Volume	2.050	03:01	
Air Blank	0.000	03:01	
Control Test	0.079	03:01	
Air Blank	0.000	03:02	
Diagnostics Check	OK	03:02	

Data Download Date/Time: 07/12/2016 12:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALVAREZ,JUAN,R
BREATH TEST OPERATOR AGENCY: Fort Pierce PD
ARRESTING OFFICER: DAVIS,KAYCEE
ARRESTING OFFICER AGENCY: Fort Pierce PD
LAST AGENCY INSPECTION: 06/17/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35213080A3
DRY GAS EXPIRY DATE: 01/01/2016
VIOLATION CODE: DUI
ARREST TIME: 20:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:24

SUBJECT'S LAST NAME: NEGRON

DATE/TIME OF TEST: 07/04/2016 21:31:22

Diagnostics Check	OK	21:28	
Air Blank	0.000	21:28	
Control Test	0.079	21:29	
Air Blank	0.000	21:29	
Subject Sample	# 1	21:30	Subject Test Refused
Breath Volume	0.000	21:30	
Air Blank	0.000	21:30	
Control Test	0.079	21:30	
Air Blank	0.000	21:31	
Diagnostics Check	OK	21:31	

Data Download Date/Time: 07/12/2016 12:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOGAN,TODD,A
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: HOGAN,TODD
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 06/17/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35213080A3
DRY GAS EXPIRY DATE: 01/01/2016
VIOLATION CODE: DUI
ARREST TIME: 21:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:53

SUBJECT'S LAST NAME: TORRES GOMEZ

DATE/TIME OF TEST: 07/04/2016 22:28:30

Diagnostics Check		OK	22:16	
Air Blank		0.000	22:17	
Control Test		0.079	22:17	
Air Blank		0.000	22:18	
Subject Sample	# 1	0.162	22:21	Volume Not Met
Breath Volume		0.871	22:21	
Air Blank		0.000	22:22	
Air Blank		0.000	22:23	
Subject Sample	# 2	0.163	22:26	Volume Not Met
Breath Volume		0.761	22:26	
Air Blank		0.000	22:27	
Control Test		0.077	22:27	
Air Blank		0.000	22:28	
Diagnostics Check		OK	22:28	

Data Download Date/Time: 07/12/2016 12:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOGAN,TODD,A
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: HOGAN,TODD
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 06/17/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35213080A3
DRY GAS EXPIRY DATE: 01/01/2016
VIOLATION CODE: DUI
ARREST TIME: 02:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: CONRAD

DATE/TIME OF TEST: 07/07/2016 03:19:14

Diagnostics Check	OK	03:12
Air Blank	0.000	03:13
Control Test	0.080	03:13
Air Blank	0.000	03:13
Subject Sample # 1	0.230	03:14
Breath Volume	1.855	03:14
Air Blank	0.000	03:15
Air Blank	0.000	03:17
Subject Sample # 2	0.229	03:17
Breath Volume	1.492	03:17
Air Blank	0.000	03:18
Control Test	0.079	03:18
Air Blank	0.000	03:19
Diagnostics Check	OK	03:19

Data Download Date/Time: 07/12/2016 12:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANTON,NICHOLAS,A
BREATH TEST OPERATOR AGENCY: Fort Pierce PD
ARRESTING OFFICER: ANTON,NICHOLAS
ARRESTING OFFICER AGENCY: Fort Pierce PD
LAST AGENCY INSPECTION: 06/17/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35213080A3
DRY GAS EXPIRY DATE: 10/05/2016
VIOLATION CODE: DUI
ARREST TIME: 20:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: DAVIDSON

DATE/TIME OF TEST: 07/07/2016 21:38:53

Diagnostics Check	OK	21:32
Air Blank	0.000	21:33
Control Test	0.079	21:33
Air Blank	0.000	21:33
Subject Sample # 1	0.000	21:34
Breath Volume	1.593	21:34
Air Blank	0.000	21:35
Air Blank	0.000	21:37
Subject Sample # 2	0.000	21:37
Breath Volume	1.503	21:37
Air Blank	0.000	21:37
Control Test	0.080	21:38
Air Blank	0.000	21:38
Diagnostics Check	OK	21:38

Data Download Date/Time: 07/12/2016 12:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WITT, TYLER, G
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: SERAPHIN, RONALD
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 06/17/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35213080A3
DRY GAS EXPIRY DATE: 01/01/2016
VIOLATION CODE: DUI
ARREST TIME: 03:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: WYMER

DATE/TIME OF TEST: 07/10/2016 04:17:56

Diagnostics Check	OK	04:09
Air Blank	0.000	04:09
Control Test	0.080	04:10
Air Blank	0.000	04:10
Subject Sample # 1	0.116	04:13
Breath Volume	1.500	04:13
Air Blank	0.000	04:14
Air Blank	0.000	04:15
Subject Sample # 2	0.119	04:16
Breath Volume	1.441	04:16
Air Blank	0.000	04:17
Control Test	0.076	04:17
Air Blank	0.000	04:17
Diagnostics Check	OK	04:17

Data Download Date/Time: 07/12/2016 12:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUTIERREZ,KRISTINA,L
BREATH TEST OPERATOR AGENCY: FHP Troop K
ARRESTING OFFICER: GUTIERREZ,KRISTINA
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 06/17/2016

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 26614080A7
DRY GAS EXPIRY DATE: 10/05/2016
VIOLATION CODE: DUI
ARREST TIME: 21:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: DE LUCK

DATE/TIME OF TEST: 07/10/2016 23:12:36

Diagnostics Check	OK	22:58	
Air Blank	0.000	22:59	
Control Test	0.079	22:59	
Air Blank	0.000	22:59	
Subject Sample	# 1 0.123	23:01	
Breath Volume	2.480	23:01	
Air Blank	0.000	23:01	
Air Blank	0.000	23:03	
Subject Sample	# 2 0.090	23:06	Volume Not Met
Breath Volume	0.363	23:06	
Air Blank	0.000	23:07	
Air Blank	0.000	23:09	
Subject Sample	# 3 0.117	23:10	
Breath Volume	1.390	23:10	
Air Blank	0.000	23:11	
Control Test	0.075	23:12	
Air Blank	0.000	23:12	
Diagnostics Check	OK	23:12	

Data Download Date/Time: 07/12/2016 12:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUTIERREZ,KRISTINA,L
BREATH TEST OPERATOR AGENCY: FHP Troop K
ARRESTING OFFICER: GUTIERREZ,KRISTINA
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 06/17/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26614080A7
DRY GAS EXPIRY DATE: 10/05/2016
VIOLATION CODE: DUI
ARREST TIME: 21:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: DE LUCK

DATE/TIME OF TEST: 07/10/2016 23:35:42

Diagnostics Check	OK	23:20	
Air Blank	0.000	23:20	
Control Test	0.077	23:21	
Air Blank	0.000	23:21	
Subject Sample	# 1 0.115	23:23	
Breath Volume	2.101	23:23	
Air Blank	0.000	23:23	
Air Blank	0.000	23:25	
Subject Sample	# 2	23:28	No Sample Provided
Breath Volume	0.000	23:28	
Air Blank	0.000	23:29	
Air Blank	0.000	23:31	
Subject Sample	# 3	23:34	Subject Test Refused
Breath Volume	0.000	23:34	
Air Blank	0.000	23:34	
Control Test	0.077	23:35	
Air Blank	0.000	23:35	
Diagnostics Check	OK	23:35	

Data Download Date/Time: 07/12/2016 12:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAYFORD,GREGG,R
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: MARGERUM,ROBERT
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 07/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A1
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 10:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:05

SUBJECT'S LAST NAME: HARVEY

DATE/TIME OF TEST: 07/15/2016 11:33:26

Diagnostics Check	OK	11:27
Air Blank	0.000	11:27
Control Test	0.080	11:27
Air Blank	0.000	11:28
Subject Sample # 1	0.000	11:29
Breath Volume	3.375	11:29
Air Blank	0.000	11:29
Air Blank	0.000	11:31
Subject Sample # 2	0.000	11:32
Breath Volume	1.472	11:32
Air Blank	0.000	11:32
Control Test	0.079	11:32
Air Blank	0.000	11:33
Diagnostics Check	OK	11:33

Data Download Date/Time: 08/25/2016 13:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUNAS,RYAN,S
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: YUNAS,RYAN
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 07/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A1
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: MALONE

DATE/TIME OF TEST: 07/17/2016 02:42:21

Diagnostics Check	OK	02:35
Air Blank	0.000	02:36
Control Test	0.080	02:36
Air Blank	0.000	02:37
Subject Sample # 1	0.197	02:37
Breath Volume	3.269	02:37
Air Blank	0.000	02:38
Air Blank	0.000	02:40
Subject Sample # 2	0.206	02:40
Breath Volume	2.191	02:40
Air Blank	0.000	02:41
Control Test	0.080	02:41
Air Blank	0.000	02:42
Diagnostics Check	OK	02:42

Data Download Date/Time: 08/25/2016 13:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAZ,ANTONIO,
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: THOMAZ,ANTONIO
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 07/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A1
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 09:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: HUERTA

DATE/TIME OF TEST: 07/18/2016 23:11:45

Diagnostics Check	OK	23:06	
Air Blank	0.000	23:06	
Control Test	0.079	23:06	
Air Blank	0.000	23:07	
Subject Sample	# 1 0.298	23:07	
Breath Volume	3.816	23:07	
Air Blank	0.000	23:08	
Air Blank	0.000	23:10	
Subject Sample	# 2 0.304	23:10	
Breath Volume	3.753	23:10	
Air Blank		23:11	RFI Detect
Air Blank	0.000	23:11	

Data Download Date/Time: 08/25/2016 13:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WRIGHT,TANYA,L
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: WALTEBURG,MARCUS
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 07/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A1
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: MONETTA

DATE/TIME OF TEST: 07/19/2016 02:06:41

Diagnostics Check	OK	02:00
Air Blank	0.000	02:00
Control Test	0.079	02:01
Air Blank	0.000	02:01
Subject Sample # 1	0.167	02:02
Breath Volume	2.300	02:02
Air Blank	0.000	02:02
Air Blank	0.000	02:04
Subject Sample # 2	0.173	02:05
Breath Volume	1.710	02:05
Air Blank	0.000	02:05
Control Test	0.079	02:06
Air Blank	0.000	02:06
Diagnostics Check	OK	02:06

Data Download Date/Time: 08/25/2016 13:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENNAN,LAURIE,B
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: BRENNAN,LAURIE
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 07/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A1
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:17

SUBJECT'S LAST NAME: ALVAREZ PEREZ

DATE/TIME OF TEST: 07/22/2016 03:45:47

Diagnostics Check	OK	03:37
Air Blank	0.000	03:38
Control Test	0.079	03:38
Air Blank	0.000	03:38
Subject Sample # 1	0.094	03:39
Breath Volume	1.636	03:39
Air Blank	0.000	03:39
Air Blank	0.000	03:41
Subject Sample # 2	0.091	03:44
Breath Volume	1.402	03:44
Air Blank	0.000	03:44
Control Test	0.079	03:45
Air Blank	0.000	03:45
Diagnostics Check	OK	03:45

Data Download Date/Time: 08/25/2016 13:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DRAWDY,LAINE,A
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: DRAWDY,LAINE
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 07/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A1
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:25

SUBJECT'S LAST NAME: TETO

DATE/TIME OF TEST: 07/23/2016 22:02:29

Diagnostics Check	OK	21:45	
Air Blank	0.000	21:46	
Control Test	0.079	21:46	
Air Blank	0.000	21:46	
Subject Sample	# 1 0.135	21:49	
Breath Volume	1.238	21:49	
Air Blank	0.000	21:50	
Air Blank	0.000	21:52	
Subject Sample	# 2 0.132	21:55	Volume Not Met
Breath Volume	0.074	21:55	
Air Blank	0.000	21:55	
Air Blank	0.000	21:57	
Subject Sample	# 3 0.119	22:00	Volume Not Met
Breath Volume	0.128	22:00	
Air Blank	0.000	22:01	
Control Test	0.078	22:01	
Air Blank	0.000	22:02	
Diagnostics Check	OK	22:02	

Data Download Date/Time: 08/25/2016 13:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,MICHAEL,L
BREATH TEST OPERATOR AGENCY: Fort Pierce PD
ARRESTING OFFICER: FRAZIER,MICHAEL
ARRESTING OFFICER AGENCY: Fort Pierce PD
LAST AGENCY INSPECTION: 07/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A1
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: BUCCI

DATE/TIME OF TEST: 07/25/2016 03:19:32

Diagnostics Check	OK	03:12
Air Blank	0.000	03:12
Control Test	0.079	03:13
Air Blank	0.000	03:13
Subject Sample # 1	0.181	03:14
Breath Volume	2.378	03:14
Air Blank	0.000	03:15
Air Blank	0.000	03:17
Subject Sample # 2	0.183	03:17
Breath Volume	2.460	03:17
Air Blank	0.000	03:18
Control Test	0.079	03:18
Air Blank	0.000	03:19
Diagnostics Check	OK	03:19

Data Download Date/Time: 08/25/2016 13:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENNAN,LAURIE,B
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: BRENNAN,LAURIE
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 07/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A1
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:14

SUBJECT'S LAST NAME: ROCK

DATE/TIME OF TEST: 07/26/2016 03:42:16

Diagnostics Check		OK	03:35
Air Blank		0.000	03:36
Control Test		0.079	03:36
Air Blank		0.000	03:37
Subject Sample	# 1	0.236	03:37
Breath Volume		2.765	03:37
Air Blank		0.000	03:38
Air Blank		0.000	03:40
Subject Sample	# 2	0.243	03:40
Breath Volume		3.253	03:40
Air Blank		0.000	03:41
Control Test		0.078	03:41
Air Blank		0.000	03:42
Diagnostics Check		OK	03:42

Data Download Date/Time: 08/25/2016 13:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARNER,TODD,A
BREATH TEST OPERATOR AGENCY: Fort Pierce PD
ARRESTING OFFICER: WARNER,TODD
ARRESTING OFFICER AGENCY: Fort Pierce PD
LAST AGENCY INSPECTION: 07/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A1
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:34

SUBJECT'S LAST NAME: JORDON

DATE/TIME OF TEST: 07/28/2016 00:02:14

Diagnostics Check	OK	23:54
Air Blank	0.000	23:54
Control Test	0.079	23:55
Air Blank	0.000	23:55
Subject Sample # 2	0.117	23:56
Breath Volume	1.652	23:56
Air Blank	0.000	23:56
Air Blank	0.000	23:58
Subject Sample # 1	0.106	00:00
Breath Volume	1.800	00:00
Air Blank	0.000	00:01
Control Test	0.078	00:01
Air Blank	0.000	00:02
Diagnostics Check	OK	00:02

Data Download Date/Time: 08/25/2016 13:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DRAWDY,LAINE,A
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: DRAWDY,LAINE
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 07/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A1
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: SANTIAGO

DATE/TIME OF TEST: 07/29/2016 00:45:29

Diagnostics Check		OK	00:33	
Air Blank		0.000	00:34	
Control Test		0.080	00:34	
Air Blank		0.000	00:35	
Subject Sample	# 1	0.168	00:38	Volume Not Met
Breath Volume		0.015	00:38	
Air Blank		0.000	00:38	
Air Blank		0.000	00:40	
Subject Sample	# 2	0.159	00:43	Volume Not Met
Breath Volume		0.000	00:43	
Air Blank		0.000	00:44	
Control Test		0.079	00:44	
Air Blank		0.000	00:45	
Diagnostics Check		OK	00:45	

Data Download Date/Time: 08/25/2016 13:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUINTO,LAWRENCE,J
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: QUINTO,LAWRENCE
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 07/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A1
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: CRUZ

DATE/TIME OF TEST: 07/29/2016 22:49:16

Diagnostics Check	OK	22:40	
Air Blank	0.000	22:41	
Control Test	0.079	22:41	
Air Blank	0.000	22:41	
Subject Sample	# 1 0.161	22:43	
Breath Volume	1.222	22:43	
Air Blank	0.000	22:44	
Air Blank	0.000	22:46	
Subject Sample	# 2	22:47	Subject Test Refused
Breath Volume	0.000	22:47	
Air Blank	0.000	22:48	
Control Test	0.078	22:48	
Air Blank	0.000	22:49	
Diagnostics Check	OK	22:49	

Data Download Date/Time: 08/25/2016 13:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENNAN,LAURIE,B
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: BRENNAN,LAURIE
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 07/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A1
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:58

SUBJECT'S LAST NAME: SPRINGER

DATE/TIME OF TEST: 07/30/2016 22:30:52

Diagnostics Check		OK	22:22	
Air Blank		0.000	22:22	
Control Test		0.079	22:23	
Air Blank		0.000	22:23	
Subject Sample	# 1	0.202	22:26	Volume Not Met
Breath Volume		0.640	22:26	
Air Blank		0.000	22:27	
Air Blank		0.000	22:29	
Subject Sample	# 2		22:29	Subject Test Refused
Breath Volume		0.000	22:29	
Air Blank		0.000	22:29	
Control Test		0.077	22:30	
Air Blank		0.000	22:30	
Diagnostics Check		OK	22:30	

Data Download Date/Time: 08/25/2016 13:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLROYD,MEGAN,G
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: TRUDEAU,RYAN
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 07/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A1
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:21

SUBJECT'S LAST NAME: SILVA

DATE/TIME OF TEST: 07/31/2016 03:56:55

Diagnostics Check	OK	03:56	
Air Blank		03:56	RFI Detect
Air Blank	0.000	03:56	

Data Download Date/Time: 08/25/2016 13:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLROYD,MEGAN,G
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: TRUDEAU,RYAN
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 07/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A1
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:21

SUBJECT'S LAST NAME: SILVA

DATE/TIME OF TEST: 07/31/2016 04:04:46

Diagnostics Check	OK	03:58
Air Blank	0.000	03:59
Control Test	0.079	03:59
Air Blank	0.000	03:59
Subject Sample # 1	0.107	04:00
Breath Volume	1.921	04:00
Air Blank	0.000	04:01
Air Blank	0.000	04:02
Subject Sample # 2	0.111	04:03
Breath Volume	2.757	04:03
Air Blank	0.000	04:03
Control Test	0.079	04:04
Air Blank	0.000	04:04
Diagnostics Check	OK	04:04

Data Download Date/Time: 08/25/2016 13:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WOODEN, D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:42

SUBJECT'S LAST NAME: CHOI

DATE/TIME OF TEST: 07/01/2016 00:11:06

Diagnostics Check	OK	00:03
Air Blank	0.000	00:03
Control Test	0.079	00:03
Air Blank	0.000	00:04
Subject Sample # 1	0.167	00:05
Breath Volume	1.414	00:05
Air Blank	0.000	00:05
Air Blank	0.000	00:07
Subject Sample # 2	0.167	00:09
Breath Volume	2.976	00:09
Air Blank	0.000	00:10
Control Test	0.078	00:10
Air Blank	0.000	00:11
Diagnostics Check	OK	00:11

Data Download Date/Time: 08/10/2016 18:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LARA, A
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 19:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:46

SUBJECT'S LAST NAME: ARANGO

DATE/TIME OF TEST: 07/01/2016 21:12:37

Diagnostics Check	OK	21:06
Air Blank	0.000	21:06
Control Test	0.079	21:07
Air Blank	0.000	21:07
Subject Sample # 1	0.168	21:08
Breath Volume	3.101	21:08
Air Blank	0.000	21:08
Air Blank	0.000	21:10
Subject Sample # 2	0.173	21:11
Breath Volume	3.136	21:11
Air Blank	0.000	21:11
Control Test	0.078	21:11
Air Blank	0.000	21:12
Diagnostics Check	OK	21:12

Data Download Date/Time: 08/10/2016 18:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: NIEUWENDAM, R
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: DICKSTEIN

DATE/TIME OF TEST: 07/02/2016 03:51:47

Diagnostics Check	OK	03:45
Air Blank	0.000	03:45
Control Test	0.079	03:46
Air Blank	0.000	03:46
Subject Sample # 1	0.025	03:47
Breath Volume	4.453	03:47
Air Blank	0.000	03:47
Air Blank	0.000	03:49
Subject Sample # 2	0.024	03:50
Breath Volume	3.878	03:50
Air Blank	0.000	03:50
Control Test	0.078	03:51
Air Blank	0.000	03:51
Diagnostics Check	OK	03:51

Data Download Date/Time: 08/10/2016 18:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MCMURPHY,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:29

SUBJECT'S LAST NAME: MOORE

DATE/TIME OF TEST: 07/02/2016 04:57:01

Diagnostics Check	OK	04:50
Air Blank	0.000	04:50
Control Test	0.079	04:51
Air Blank	0.000	04:51
Subject Sample # 1	0.169	04:52
Breath Volume	2.113	04:52
Air Blank	0.000	04:53
Air Blank	0.000	04:54
Subject Sample # 2	0.178	04:55
Breath Volume	3.277	04:55
Air Blank	0.000	04:56
Control Test	0.077	04:56
Air Blank	0.000	04:56
Diagnostics Check	OK	04:57

Data Download Date/Time: 08/10/2016 18:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,REBECCA,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: AGUAYO,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 17:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:56

SUBJECT'S LAST NAME: TITTLE Y RIVERA

DATE/TIME OF TEST: 07/02/2016 18:29:21

Diagnostics Check	OK	18:22
Air Blank	0.000	18:22
Control Test	0.079	18:22
Air Blank	0.000	18:23
Subject Sample # 1	0.323	18:24
Breath Volume	3.527	18:24
Air Blank	0.000	18:25
Air Blank	0.000	18:27
Subject Sample # 2	0.328	18:27
Breath Volume	3.535	18:27
Air Blank	0.000	18:28
Control Test	0.078	18:28
Air Blank	0.000	18:29
Diagnostics Check	OK	18:29

Data Download Date/Time: 08/10/2016 18:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GONZALEZ,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:18

SUBJECT'S LAST NAME: TREADWAY

DATE/TIME OF TEST: 07/03/2016 01:50:35

Diagnostics Check	OK	01:43
Air Blank	0.000	01:44
Control Test	0.079	01:44
Air Blank	0.000	01:45
Subject Sample # 1	0.074	01:46
Breath Volume	2.476	01:46
Air Blank	0.000	01:46
Air Blank	0.000	01:48
Subject Sample # 2	0.066	01:49
Breath Volume	1.519	01:49
Air Blank	0.000	01:49
Control Test	0.078	01:49
Air Blank	0.000	01:50
Diagnostics Check	OK	01:50

Data Download Date/Time: 08/10/2016 18:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK,CASEY,S
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: FLORES

DATE/TIME OF TEST: 07/03/2016 03:34:24

Diagnostics Check	OK	03:28
Air Blank	0.000	03:28
Control Test	0.079	03:29
Air Blank	0.000	03:29
Subject Sample # 1	0.159	03:29
Breath Volume	1.398	03:29
Air Blank	0.000	03:30
Air Blank	0.000	03:32
Subject Sample # 2	0.158	03:32
Breath Volume	1.546	03:32
Air Blank	0.000	03:33
Control Test	0.079	03:33
Air Blank	0.000	03:34
Diagnostics Check	OK	03:34

Data Download Date/Time: 08/10/2016 18:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNYDER,MELONIE,E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARFIELD-CRAIG,C
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 10:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 11:32

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 07/03/2016 11:59:22

Diagnostics Check	OK	11:53
Air Blank	0.000	11:53
Control Test	0.080	11:53
Air Blank	0.000	11:54
Subject Sample # 1	0.000	11:55
Breath Volume	2.976	11:55
Air Blank	0.000	11:55
Air Blank	0.000	11:57
Subject Sample # 2	0.000	11:58
Breath Volume	2.382	11:58
Air Blank	0.000	11:58
Control Test	0.079	11:58
Air Blank	0.000	11:59
Diagnostics Check	OK	11:59

Data Download Date/Time: 08/10/2016 18:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,REBECCA,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FRENCH,E
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 15:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:51

SUBJECT'S LAST NAME: ALMAZRUA

DATE/TIME OF TEST: 07/03/2016 16:19:56

Diagnostics Check	OK	16:13
Air Blank	0.000	16:14
Control Test	0.079	16:14
Air Blank	0.000	16:15
Subject Sample # 1	0.000	16:15
Breath Volume	2.968	16:15
Air Blank	0.000	16:16
Air Blank	0.000	16:18
Subject Sample # 2	0.000	16:18
Breath Volume	2.847	16:18
Air Blank	0.000	16:19
Control Test	0.079	16:19
Air Blank	0.000	16:19
Diagnostics Check	OK	16:19

Data Download Date/Time: 08/10/2016 18:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:12

SUBJECT'S LAST NAME: HEFLIN

DATE/TIME OF TEST: 07/03/2016 23:45:21

Diagnostics Check		OK	23:33	
Air Blank		0.000	23:34	
Control Test		0.079	23:34	
Air Blank		0.000	23:34	
Subject Sample	# 1	0.183	23:38	Volume Not Met
Breath Volume		0.554	23:38	
Air Blank		0.000	23:38	
Air Blank		0.000	23:40	
Subject Sample	# 2	0.190	23:43	Volume Not Met
Breath Volume		0.664	23:43	
Air Blank		0.000	23:44	
Control Test		0.077	23:44	
Air Blank		0.000	23:45	
Diagnostics Check		OK	23:45	

Data Download Date/Time: 08/10/2016 18:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:12

SUBJECT'S LAST NAME: HEFLIN

DATE/TIME OF TEST: 07/04/2016 00:00:39

Diagnostics Check		OK	23:49	
Air Blank		0.000	23:49	
Control Test		0.078	23:49	
Air Blank		0.000	23:50	
Subject Sample	# 1	0.168	23:53	Volume Not Met
Breath Volume		0.042	23:53	
Air Blank		0.000	23:54	
Air Blank		0.000	23:55	
Subject Sample	# 2	0.166	23:59	Volume Not Met
Breath Volume		0.132	23:59	
Air Blank		0.000	23:59	
Control Test		0.077	00:00	
Air Blank		0.000	00:00	
Diagnostics Check		OK	00:00	

Data Download Date/Time: 08/10/2016 18:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BLANCHARD,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:26

SUBJECT'S LAST NAME: EDWARDS III

DATE/TIME OF TEST: 07/04/2016 01:50:46

Diagnostics Check	OK	01:46	
Air Blank	0.000	01:47	
Control Test	0.079	01:47	
Air Blank	0.000	01:47	
Subject Sample	# 1	01:49	Subject Test Refused
Breath Volume	0.000	01:49	
Air Blank	0.000	01:49	
Control Test	0.078	01:50	
Air Blank	0.000	01:50	
Diagnostics Check	OK	01:50	

Data Download Date/Time: 08/10/2016 18:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SIMPKINS,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:24

SUBJECT'S LAST NAME: SANTOS JR

DATE/TIME OF TEST: 07/04/2016 04:52:52

Diagnostics Check	OK	04:46
Air Blank	0.000	04:46
Control Test	0.079	04:47
Air Blank	0.000	04:47
Subject Sample # 1	0.153	04:48
Breath Volume	3.609	04:48
Air Blank	0.000	04:48
Air Blank	0.000	04:50
Subject Sample # 2	0.149	04:51
Breath Volume	2.968	04:51
Air Blank	0.000	04:51
Control Test	0.079	04:52
Air Blank	0.000	04:52
Diagnostics Check	OK	04:52

Data Download Date/Time: 08/10/2016 18:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DUNCAN,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:37

SUBJECT'S LAST NAME: BAILEY

DATE/TIME OF TEST: 07/04/2016 05:09:02

Diagnostics Check	OK	04:59
Air Blank	0.000	05:00
Control Test	0.078	05:00
Air Blank	0.000	05:00
Subject Sample # 1	0.115	05:03
Breath Volume	1.343	05:03
Air Blank	0.000	05:03
Air Blank	0.000	05:05
Subject Sample # 2	0.110	05:07
Breath Volume	1.207	05:07
Air Blank	0.000	05:08
Control Test	0.078	05:08
Air Blank	0.000	05:08
Diagnostics Check	OK	05:09

Data Download Date/Time: 08/10/2016 18:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,REBECCA,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WOODEN,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 17:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:00

SUBJECT'S LAST NAME: KUSSMAN

DATE/TIME OF TEST: 07/04/2016 18:28:41

Diagnostics Check	OK	18:20
Air Blank	0.000	18:21
Control Test	0.079	18:21
Air Blank	0.000	18:22
Subject Sample # 1	0.184	18:23
Breath Volume	1.550	18:23
Air Blank	0.000	18:24
Air Blank	0.000	18:26
Subject Sample # 2	0.185	18:27
Breath Volume	1.445	18:27
Air Blank	0.000	18:27
Control Test	0.077	18:28
Air Blank	0.000	18:28
Diagnostics Check	OK	18:28

Data Download Date/Time: 08/10/2016 18:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:48

SUBJECT'S LAST NAME: LANG JR

DATE/TIME OF TEST: 07/05/2016 02:17:35

Diagnostics Check	OK	02:08
Air Blank	0.000	02:08
Control Test	0.079	02:09
Air Blank	0.000	02:09
Subject Sample # 1	0.155	02:12
Breath Volume	1.691	02:12
Air Blank	0.000	02:13
Air Blank	0.000	02:15
Subject Sample # 2	0.159	02:16
Breath Volume	1.519	02:16
Air Blank	0.000	02:16
Control Test	0.078	02:16
Air Blank	0.000	02:17
Diagnostics Check	OK	02:17

Data Download Date/Time: 08/10/2016 18:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: CRAGG, S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:24

SUBJECT'S LAST NAME: MCCARRELL

DATE/TIME OF TEST: 07/06/2016 01:51:41

Diagnostics Check		OK	01:45
Air Blank		0.000	01:45
Control Test		0.079	01:46
Air Blank		0.000	01:46
Subject Sample	# 1	0.148	01:47
Breath Volume		1.484	01:47
Air Blank		0.000	01:47
Air Blank		0.000	01:49
Subject Sample	# 2	0.154	01:50
Breath Volume		1.109	01:50
Air Blank		0.000	01:50
Control Test		0.079	01:51
Air Blank		0.000	01:51
Diagnostics Check		OK	01:51

Data Download Date/Time: 08/10/2016 18:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY,PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DUNCAN,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: KOGGE

DATE/TIME OF TEST: 07/07/2016 02:43:02

Diagnostics Check	OK	02:36
Air Blank	0.000	02:36
Control Test	0.079	02:37
Air Blank	0.000	02:37
Subject Sample # 1	0.187	02:38
Breath Volume	3.285	02:38
Air Blank	0.000	02:38
Air Blank	0.000	02:40
Subject Sample # 2	0.184	02:41
Breath Volume	3.171	02:41
Air Blank	0.000	02:42
Control Test	0.078	02:42
Air Blank	0.000	02:42
Diagnostics Check	OK	02:43

Data Download Date/Time: 08/10/2016 18:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MALONE,C
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: VICENTE

DATE/TIME OF TEST: 07/07/2016 23:42:02

Diagnostics Check	OK	23:35
Air Blank	0.000	23:36
Control Test	0.079	23:36
Air Blank	0.000	23:36
Subject Sample # 1	0.078	23:37
Breath Volume	2.433	23:37
Air Blank	0.000	23:38
Air Blank	0.000	23:40
Subject Sample # 2	0.081	23:40
Breath Volume	2.644	23:40
Air Blank	0.000	23:41
Control Test	0.079	23:41
Air Blank	0.000	23:41
Diagnostics Check	OK	23:42

Data Download Date/Time: 08/10/2016 18:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HALL,K
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:53

SUBJECT'S LAST NAME: CORTES

DATE/TIME OF TEST: 07/09/2016 00:20:42

Diagnostics Check	OK	00:14
Air Blank	0.000	00:15
Control Test	0.079	00:15
Air Blank	0.000	00:15
Subject Sample # 1	0.000	00:16
Breath Volume	1.386	00:16
Air Blank	0.000	00:16
Air Blank	0.000	00:18
Subject Sample # 2	0.000	00:19
Breath Volume	2.054	00:19
Air Blank	0.000	00:19
Control Test	0.078	00:20
Air Blank	0.000	00:20
Diagnostics Check	OK	00:20

Data Download Date/Time: 08/10/2016 18:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DAWSY,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:23

SUBJECT'S LAST NAME: WANAMAKER

DATE/TIME OF TEST: 07/09/2016 01:48:15

Diagnostics Check	OK	01:43	
Air Blank	0.000	01:43	
Control Test	0.078	01:44	
Air Blank	0.000	01:44	
Subject Sample	# 1	01:46	Subject Test Refused
Breath Volume	0.000	01:46	
Air Blank	0.000	01:47	
Control Test	0.079	01:47	
Air Blank	0.000	01:48	
Diagnostics Check	OK	01:48	

Data Download Date/Time: 08/10/2016 18:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK,CASEY,S
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SIMPKINS,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:08

SUBJECT'S LAST NAME: ALMARAZ-RAMIREZ

DATE/TIME OF TEST: 07/09/2016 04:34:43

Diagnostics Check	OK	04:28
Air Blank	0.000	04:28
Control Test	0.078	04:29
Air Blank	0.000	04:29
Subject Sample # 1	0.168	04:30
Breath Volume	1.933	04:30
Air Blank	0.000	04:30
Air Blank	0.000	04:32
Subject Sample # 2	0.170	04:33
Breath Volume	2.460	04:33
Air Blank	0.000	04:33
Control Test	0.078	04:34
Air Blank	0.000	04:34
Diagnostics Check	OK	04:34

Data Download Date/Time: 08/10/2016 18:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HALL,K
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:37

SUBJECT'S LAST NAME: HAMILTON

DATE/TIME OF TEST: 07/09/2016 06:06:08

Diagnostics Check	OK	05:57
Air Blank	0.000	05:58
Control Test	0.079	05:58
Air Blank	0.000	05:59
Subject Sample # 1	0.105	06:01
Breath Volume	2.191	06:01
Air Blank	0.000	06:01
Air Blank	0.000	06:03
Subject Sample # 2	0.098	06:04
Breath Volume	1.214	06:04
Air Blank	0.000	06:05
Control Test	0.078	06:05
Air Blank	0.000	06:06
Diagnostics Check	OK	06:06

Data Download Date/Time: 08/10/2016 18:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMMONS,ANGELA,V
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: NAYLOR,I
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 06:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:20

SUBJECT'S LAST NAME: GOMEZ

DATE/TIME OF TEST: 07/09/2016 07:46:39

Diagnostics Check	OK	07:40	
Air Blank	0.000	07:40	
Control Test	0.079	07:41	
Air Blank	0.000	07:41	
Subject Sample	# 1 0.111	07:42	
Breath Volume	1.156	07:42	
Air Blank	0.000	07:42	
Air Blank	0.000	07:44	
Subject Sample	# 2	07:45	Subject Test Refused
Breath Volume	0.000	07:45	
Air Blank	0.000	07:45	
Control Test	0.078	07:46	
Air Blank	0.000	07:46	
Diagnostics Check	OK	07:46	

Data Download Date/Time: 08/10/2016 18:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMMONS,ANGELA,V
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: NAYLOR,I
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 06:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:47

SUBJECT'S LAST NAME: GOMEZ

DATE/TIME OF TEST: 07/09/2016 08:15:04

Diagnostics Check	OK	08:08
Air Blank	0.000	08:08
Control Test	0.079	08:08
Air Blank	0.000	08:09
Subject Sample # 1	0.102	08:10
Breath Volume	1.257	08:10
Air Blank	0.000	08:11
Air Blank	0.000	08:13
Subject Sample # 2	0.107	08:13
Breath Volume	2.406	08:13
Air Blank	0.000	08:14
Control Test	0.078	08:14
Air Blank	0.000	08:14
Diagnostics Check	OK	08:15

Data Download Date/Time: 08/10/2016 18:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HERNANDEZ, A
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:09

SUBJECT'S LAST NAME: VOGLER

DATE/TIME OF TEST: 07/10/2016 01:37:47

Diagnostics Check	OK	01:30
Air Blank	0.000	01:30
Control Test	0.079	01:31
Air Blank	0.000	01:31
Subject Sample # 1	0.155	01:32
Breath Volume	2.113	01:32
Air Blank	0.000	01:33
Air Blank	0.000	01:35
Subject Sample # 2	0.155	01:36
Breath Volume	1.250	01:36
Air Blank	0.000	01:36
Control Test	0.078	01:37
Air Blank	0.000	01:37
Diagnostics Check	OK	01:37

Data Download Date/Time: 08/10/2016 18:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY,PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DAWSY,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:51

SUBJECT'S LAST NAME: ZOET

DATE/TIME OF TEST: 07/10/2016 03:18:47

Diagnostics Check	OK	03:11
Air Blank	0.000	03:11
Control Test	0.079	03:12
Air Blank	0.000	03:12
Subject Sample # 1	0.172	03:13
Breath Volume	2.046	03:13
Air Blank	0.000	03:14
Air Blank	0.000	03:16
Subject Sample # 2	0.186	03:17
Breath Volume	1.960	03:17
Air Blank	0.000	03:17
Control Test	0.078	03:18
Air Blank	0.000	03:18
Diagnostics Check	OK	03:18

Data Download Date/Time: 08/10/2016 18:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRABAND,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: CHURCHILL

DATE/TIME OF TEST: 07/10/2016 03:53:58

Diagnostics Check	OK	03:45
Air Blank	0.000	03:46
Control Test	0.078	03:46
Air Blank	0.000	03:46
Subject Sample # 1	0.131	03:48
Breath Volume	1.375	03:48
Air Blank	0.000	03:49
Air Blank	0.000	03:51
Subject Sample # 2	0.141	03:52
Breath Volume	2.070	03:52
Air Blank	0.000	03:53
Control Test	0.078	03:53
Air Blank	0.000	03:53
Diagnostics Check	OK	03:53

Data Download Date/Time: 08/10/2016 18:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SULLIVAN,T
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:53

SUBJECT'S LAST NAME: LIMON

DATE/TIME OF TEST: 07/10/2016 05:20:12

Diagnostics Check	OK	05:13
Air Blank	0.000	05:14
Control Test	0.080	05:14
Air Blank	0.000	05:14
Subject Sample # 1	0.000	05:15
Breath Volume	2.769	05:15
Air Blank	0.000	05:16
Air Blank	0.000	05:18
Subject Sample # 2	0.000	05:18
Breath Volume	3.140	05:18
Air Blank	0.000	05:19
Control Test	0.078	05:19
Air Blank	0.000	05:20
Diagnostics Check	OK	05:20

Data Download Date/Time: 08/10/2016 18:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARR, BARBARA, H
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DALEY, N
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 07:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:26

SUBJECT'S LAST NAME: LYNCH

DATE/TIME OF TEST: 07/10/2016 08:53:54

Diagnostics Check	OK	08:47
Air Blank	0.000	08:47
Control Test	0.079	08:48
Air Blank	0.000	08:48
Subject Sample # 1	0.130	08:49
Breath Volume	2.210	08:49
Air Blank	0.000	08:49
Air Blank	0.000	08:51
Subject Sample # 2	0.133	08:52
Breath Volume	2.320	08:52
Air Blank	0.000	08:52
Control Test	0.079	08:53
Air Blank	0.000	08:53
Diagnostics Check	OK	08:53

Data Download Date/Time: 08/10/2016 18:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 07/11/2016 03:17:32

Diagnostics Check		OK	03:11
Air Blank		0.000	03:11
Control Test		0.079	03:11
Air Blank		0.000	03:12
Subject Sample	# 1	0.117	03:12
Breath Volume		3.285	03:12
Air Blank		0.000	03:13
Air Blank		0.000	03:15
Subject Sample	# 2	0.118	03:16
Breath Volume		3.781	03:16
Air Blank		0.000	03:16
Control Test		0.079	03:16
Air Blank		0.000	03:17
Diagnostics Check		OK	03:17

Data Download Date/Time: 08/10/2016 18:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MALONE, C
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:02

SUBJECT'S LAST NAME: VIGLIATURA

DATE/TIME OF TEST: 07/12/2016 01:28:53

Diagnostics Check	OK	01:22
Air Blank	0.000	01:22
Control Test	0.079	01:23
Air Blank	0.000	01:23
Subject Sample # 1	0.000	01:24
Breath Volume	3.031	01:24
Air Blank	0.000	01:24
Air Blank	0.000	01:26
Subject Sample # 2	0.000	01:27
Breath Volume	3.593	01:27
Air Blank	0.000	01:27
Control Test	0.079	01:28
Air Blank	0.000	01:28
Diagnostics Check	OK	01:28

Data Download Date/Time: 08/10/2016 18:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEIRCE, HAROLD, F
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: FIELDS, CARISSA
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/19/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A2
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 13:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:16

SUBJECT'S LAST NAME: MULDER-ALGOOD

DATE/TIME OF TEST: 07/31/2016 14:42:57

Diagnostics Check	OK	14:36
Air Blank	0.000	14:37
Control Test	0.079	14:37
Air Blank	0.000	14:38
Subject Sample # 1	0.000	14:38
Breath Volume	1.500	14:38
Air Blank	0.000	14:39
Air Blank	0.000	14:41
Subject Sample # 2	0.000	14:41
Breath Volume	1.617	14:41
Air Blank	0.000	14:42
Control Test	0.079	14:42
Air Blank	0.000	14:42
Diagnostics Check	OK	14:42

Data Download Date/Time: 08/19/2016 05:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,GARY,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: MILLER,GARY
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 06/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 83550
DRY GAS EXPIRY DATE: 09/26/2018
VIOLATION CODE: DUI
ARREST TIME: 17:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:13

SUBJECT'S LAST NAME: CALHOUN

DATE/TIME OF TEST: 07/01/2016 18:41:08

Diagnostics Check	OK	18:33	
Air Blank	0.000	18:34	
Control Test	0.078	18:34	
Air Blank	0.000	18:35	
Subject Sample # 1	0.149	18:35	
Breath Volume	1.632	18:35	
Air Blank	0.000	18:36	
Air Blank	0.000	18:38	
Subject Sample # 2	0.158	18:39	
Breath Volume	2.714	18:39	
Air Blank		18:40	Purge Fail
Air Blank	0.000	18:41	

Data Download Date/Time: 07/21/2016 12:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,GARY,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: MILLER,GARY
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 06/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 83550
DRY GAS EXPIRY DATE: 09/26/2018
VIOLATION CODE: DUI
ARREST TIME: 17:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:13

SUBJECT'S LAST NAME: CALHOUN

DATE/TIME OF TEST: 07/01/2016 18:52:15

Diagnostics Check	OK	18:43
Air Blank	0.000	18:44
Control Test	0.078	18:44
Air Blank	0.000	18:44
Subject Sample # 1	0.150	18:45
Breath Volume	1.835	18:45
Air Blank	0.000	18:46
Air Blank	0.000	18:48
Subject Sample # 2	0.148	18:50
Breath Volume	2.167	18:50
Air Blank	0.000	18:51
Control Test	0.077	18:51
Air Blank	0.000	18:52
Diagnostics Check	OK	18:52

Data Download Date/Time: 07/21/2016 12:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNOW,DONALD,R
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: SNOW,DONALD
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 06/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 83550
DRY GAS EXPIRY DATE: 09/26/2018
VIOLATION CODE: DUI
ARREST TIME: 18:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:22

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 07/03/2016 18:54:49

Diagnostics Check	OK	18:50	
Air Blank	0.000	18:51	
Control Test	0.078	18:51	
Air Blank	0.000	18:52	
Subject Sample	# 1	18:53	Subject Test Refused
Breath Volume	0.000	18:53	
Air Blank	0.000	18:53	
Control Test	0.079	18:54	
Air Blank	0.000	18:54	
Diagnostics Check	OK	18:54	

Data Download Date/Time: 07/21/2016 12:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HULL,ROBERT,E
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: HULL,ROBERT
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 06/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1662060
DRY GAS EXPIRY DATE: 04/18/2017
VIOLATION CODE: DUI
ARREST TIME: 19:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:10

SUBJECT'S LAST NAME: LINDSEY

DATE/TIME OF TEST: 07/03/2016 20:01:33

Diagnostics Check	OK	19:53	
Air Blank	0.000	19:54	
Control Test	0.079	19:54	
Air Blank	0.000	19:54	
Subject Sample	# 1 0.169	19:57	
Breath Volume	2.464	19:57	
Air Blank	0.000	19:58	
Air Blank	0.000	20:00	
Subject Sample	# 2 0.188	20:00	
Breath Volume	2.636	20:00	
Air Blank		20:01	RFI Detect
Air Blank		20:01	RFI Detect

Data Download Date/Time: 07/21/2016 12:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HULL,ROBERT,E
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: HULL,ROBERT
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 06/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1662060
DRY GAS EXPIRY DATE: 04/18/2017
VIOLATION CODE: DUI
ARREST TIME: 19:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:10

SUBJECT'S LAST NAME: LINDSEY

DATE/TIME OF TEST: 07/03/2016 20:13:57

Diagnostics Check	OK	20:07
Air Blank	0.000	20:07
Control Test	0.078	20:07
Air Blank	0.000	20:08
Subject Sample # 1	0.167	20:09
Breath Volume	2.230	20:09
Air Blank	0.000	20:09
Air Blank	0.000	20:11
Subject Sample # 2	0.176	20:12
Breath Volume	2.332	20:12
Air Blank	0.000	20:13
Control Test	0.079	20:13
Air Blank	0.000	20:13
Diagnostics Check	OK	20:13

Data Download Date/Time: 07/21/2016 12:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALLER,JESSICA,G
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: RAINES,ELDON
ARRESTING OFFICER AGENCY: Dunnellon PD
LAST AGENCY INSPECTION: 06/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 83550
DRY GAS EXPIRY DATE: 09/26/2018
VIOLATION CODE: DUI
ARREST TIME: 19:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:17

SUBJECT'S LAST NAME: SIMON

DATE/TIME OF TEST: 07/03/2016 20:53:22

Diagnostics Check		OK	20:37	
Air Blank		0.000	20:38	
Control Test		0.078	20:38	
Air Blank		0.000	20:38	
Subject Sample	# 1	0.237	20:41	Volume Not Met
Breath Volume		0.480	20:41	
Air Blank		0.000	20:42	
Air Blank		0.000	20:44	
Subject Sample	# 2	0.250	20:46	
Breath Volume		1.589	20:46	
Air Blank		0.000	20:47	
Air Blank		0.000	20:48	
Subject Sample	# 3	0.236	20:51	
Breath Volume		1.312	20:51	
Air Blank		0.000	20:52	
Control Test		0.077	20:52	
Air Blank		0.000	20:53	
Diagnostics Check		OK	20:53	

Data Download Date/Time: 07/21/2016 12:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BALL,CODY,W
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: BALL,CODY
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 06/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 83550
DRY GAS EXPIRY DATE: 09/26/2018
VIOLATION CODE: DUI
ARREST TIME: 20:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:05

SUBJECT'S LAST NAME: PIPER

DATE/TIME OF TEST: 07/03/2016 21:13:57

Diagnostics Check	OK	21:07
Air Blank	0.000	21:07
Control Test	0.077	21:08
Air Blank	0.000	21:08
Subject Sample # 1	0.030	21:09
Breath Volume	1.589	21:09
Air Blank	0.000	21:09
Air Blank	0.000	21:11
Subject Sample # 2	0.028	21:12
Breath Volume	1.132	21:12
Air Blank	0.000	21:13
Control Test	0.079	21:13
Air Blank	0.000	21:13
Diagnostics Check	OK	21:13

Data Download Date/Time: 07/21/2016 12:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERNANDEZ,ALAIN,
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: FERNANDEZ,ALAIN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 06/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 83550
DRY GAS EXPIRY DATE: 09/26/2018
VIOLATION CODE: DUI
ARREST TIME: 20:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:35

SUBJECT'S LAST NAME: CASSELMAN

DATE/TIME OF TEST: 07/04/2016 21:58:01

Diagnostics Check	OK	21:55	
Air Blank	0.000	21:55	
Control Test	0.078	21:55	
Air Blank	0.000	21:56	
Subject Sample	# 1	21:56	Subject Test Refused
Breath Volume	0.000	21:56	
Air Blank	0.000	21:57	
Control Test	0.079	21:57	
Air Blank	0.000	21:57	
Diagnostics Check	OK	21:58	

Data Download Date/Time: 07/21/2016 12:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PETERS,CHRISTINE,J
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: WINKLER,TODD
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 06/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 83550
DRY GAS EXPIRY DATE: 09/26/2018
VIOLATION CODE: DUI
ARREST TIME: 10:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:50

SUBJECT'S LAST NAME: GREEN

DATE/TIME OF TEST: 07/06/2016 11:21:31

Diagnostics Check	OK	11:16	
Air Blank	0.000	11:16	
Control Test	0.079	11:16	
Air Blank	0.000	11:17	
Subject Sample # 1	0.000	11:18	
Breath Volume	1.855	11:18	
Air Blank	0.000	11:18	
Air Blank		11:20	Improper Sample
Air Blank	0.000	11:20	
Control Test	0.080	11:20	
Air Blank	0.000	11:21	
Diagnostics Check	OK	11:21	

Data Download Date/Time: 07/21/2016 12:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PETERS,CHRISTINE,J
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: WINKLER,TODD
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 06/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 83550
DRY GAS EXPIRY DATE: 09/26/2018
VIOLATION CODE: DUI
ARREST TIME: 10:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:50

SUBJECT'S LAST NAME: GREEN

DATE/TIME OF TEST: 07/06/2016 11:30:49

Diagnostics Check	OK	11:23
Air Blank	0.000	11:24
Control Test	0.079	11:24
Air Blank	0.000	11:25
Subject Sample # 1	0.000	11:25
Breath Volume	2.148	11:25
Air Blank	0.000	11:26
Air Blank	0.000	11:28
Subject Sample # 2	0.000	11:29
Breath Volume	3.171	11:29
Air Blank	0.000	11:29
Control Test	0.078	11:30
Air Blank	0.000	11:30
Diagnostics Check	OK	11:30

Data Download Date/Time: 07/21/2016 12:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERNANDEZ,ALAIN,
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: FERNANDEZ,ALAIN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 06/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 83550
DRY GAS EXPIRY DATE: 09/26/2018
VIOLATION CODE: DUI
ARREST TIME: 03:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: BRUNELLE

DATE/TIME OF TEST: 07/10/2016 03:57:33

Diagnostics Check		OK	03:51
Air Blank		0.000	03:51
Control Test		0.079	03:51
Air Blank		0.000	03:52
Subject Sample	# 1	0.239	03:52
Breath Volume		4.101	03:52
Air Blank		0.000	03:53
Air Blank		0.000	03:55
Subject Sample	# 2	0.237	03:55
Breath Volume		3.453	03:55
Air Blank		0.000	03:56
Control Test		0.079	03:56
Air Blank		0.000	03:57
Diagnostics Check		OK	03:57

Data Download Date/Time: 07/21/2016 12:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOBOSIEWICZ,ANTHONY,N
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: DOBOSIEWICZ,ANTHONY
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 06/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 83550
DRY GAS EXPIRY DATE: 09/26/2018
VIOLATION CODE: DUI
ARREST TIME: 22:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:05

SUBJECT'S LAST NAME: TIER

DATE/TIME OF TEST: 07/11/2016 23:33:35

Diagnostics Check	OK	23:26
Air Blank	0.000	23:26
Control Test	0.079	23:26
Air Blank	0.000	23:27
Subject Sample # 1	0.168	23:28
Breath Volume	1.707	23:28
Air Blank	0.000	23:28
Air Blank	0.000	23:30
Subject Sample # 2	0.164	23:32
Breath Volume	2.222	23:32
Air Blank	0.000	23:32
Control Test	0.079	23:32
Air Blank	0.000	23:33
Diagnostics Check	OK	23:33

Data Download Date/Time: 07/21/2016 12:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,GARY,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: MILLER,GARY
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 06/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 83550
DRY GAS EXPIRY DATE: 09/26/2018
VIOLATION CODE: DUI
ARREST TIME: 01:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:46

SUBJECT'S LAST NAME: MONTHONY

DATE/TIME OF TEST: 07/16/2016 02:12:45

Diagnostics Check	OK	02:06
Air Blank	0.000	02:06
Control Test	0.079	02:07
Air Blank	0.000	02:07
Subject Sample # 1	0.212	02:08
Breath Volume	1.640	02:08
Air Blank	0.000	02:08
Air Blank	0.000	02:10
Subject Sample # 2	0.217	02:11
Breath Volume	1.664	02:11
Air Blank	0.000	02:11
Control Test	0.079	02:12
Air Blank	0.000	02:12
Diagnostics Check	OK	02:12

Data Download Date/Time: 07/21/2016 12:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,GARY,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: MILLER,GARY
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 06/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 83550
DRY GAS EXPIRY DATE: 09/26/2018
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: TURO

DATE/TIME OF TEST: 07/20/2016 04:03:50

Diagnostics Check	OK	03:57
Air Blank	0.000	03:57
Control Test	0.078	03:58
Air Blank	0.000	03:58
Subject Sample # 1	0.190	03:59
Breath Volume	1.851	03:59
Air Blank	0.000	03:59
Air Blank	0.000	04:01
Subject Sample # 2	0.184	04:02
Breath Volume	1.753	04:02
Air Blank	0.000	04:02
Control Test	0.078	04:03
Air Blank	0.000	04:03
Diagnostics Check	OK	04:03

Data Download Date/Time: 07/21/2016 12:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO,I,
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: SMITH,B
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 06/17/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: SINGER

DATE/TIME OF TEST: 07/17/2016 02:26:36

Diagnostics Check	OK	02:20
Air Blank	0.000	02:20
Control Test	0.081	02:21
Air Blank	0.000	02:21
Subject Sample # 1	0.149	02:21
Breath Volume	1.894	02:21
Air Blank	0.000	02:22
Air Blank	0.000	02:24
Subject Sample # 2	0.136	02:25
Breath Volume	1.609	02:25
Air Blank	0.000	02:25
Control Test	0.080	02:25
Air Blank	0.000	02:26
Diagnostics Check	OK	02:26

Data Download Date/Time: 07/19/2016 08:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARE,GLENN,A
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: HANSEN,ROBERT
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 07/31/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28814080A3
DRY GAS EXPIRY DATE: 11/05/2016
VIOLATION CODE: DUI
ARREST TIME: 19:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:41

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 07/31/2016 20:27:16

Diagnostics Check	OK	20:22	
Air Blank	0.000	20:22	
Control Test	0.077	20:22	
Air Blank	0.000	20:23	
Subject Sample	# 1	20:25	Subject Test Refused
Breath Volume	0.000	20:25	
Air Blank	0.000	20:26	
Control Test	0.077	20:26	
Air Blank	0.000	20:27	
Diagnostics Check	OK	20:27	

Data Download Date/Time: 08/28/2016 21:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSON,BRENDA,R
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: DEARMOND,MICHAEL
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 06/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28814080A3
DRY GAS EXPIRY DATE: 11/05/2016
VIOLATION CODE: DUI
ARREST TIME: 15:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:48

SUBJECT'S LAST NAME: KRAMER

DATE/TIME OF TEST: 07/03/2016 16:29:45

Diagnostics Check	OK	16:25	
Air Blank	0.000	16:25	
Control Test	0.076	16:26	
Air Blank	0.000	16:26	
Subject Sample	# 1 0.231	16:27	
Breath Volume	1.824	16:27	
Air Blank		16:28	RFI Detect
Air Blank	0.000	16:28	
Control Test	0.076	16:29	
Air Blank	0.000	16:29	
Diagnostics Check	OK	16:29	

Data Download Date/Time: 07/31/2016 14:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSON,BRENDA,R
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: DEARMOND,MICHEAL
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 06/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28814080A3
DRY GAS EXPIRY DATE: 11/05/2016
VIOLATION CODE: DUI
ARREST TIME: 15:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:48

SUBJECT'S LAST NAME: KRAMER

DATE/TIME OF TEST: 07/03/2016 16:39:51

Diagnostics Check	OK	16:32
Air Blank	0.000	16:33
Control Test	0.076	16:33
Air Blank	0.000	16:34
Subject Sample # 1	0.233	16:34
Breath Volume	1.664	16:34
Air Blank	0.000	16:35
Air Blank	0.000	16:37
Subject Sample # 2	0.237	16:38
Breath Volume	1.910	16:38
Air Blank	0.000	16:38
Control Test	0.075	16:39
Air Blank	0.000	16:39
Diagnostics Check	OK	16:39

Data Download Date/Time: 07/31/2016 14:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RANDALL,JUDY,C
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: RAY,ALBERT
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 06/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28814080A3
DRY GAS EXPIRY DATE: 11/05/2016
VIOLATION CODE: DUI
ARREST TIME: 23:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:46

SUBJECT'S LAST NAME: BLECZINSKI

DATE/TIME OF TEST: 07/05/2016 00:23:28

Diagnostics Check	OK	00:11	
Air Blank	0.000	00:12	
Control Test	0.076	00:12	
Air Blank	0.000	00:13	
Subject Sample	# 1 0.149	00:16	Volume Not Met
Breath Volume	0.402	00:16	
Air Blank	0.000	00:17	
Air Blank	0.000	00:18	
Subject Sample	# 2	00:21	Subject Test Refused
Breath Volume	0.523	00:21	
Air Blank	0.000	00:22	
Control Test	0.075	00:22	
Air Blank	0.000	00:23	
Diagnostics Check	OK	00:23	

Data Download Date/Time: 07/31/2016 14:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REID,JAMES,D
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: RAY,A
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 06/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28814080A3
DRY GAS EXPIRY DATE: 11/05/2016
VIOLATION CODE: DUI
ARREST TIME: 23:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: BLECZINSKI

DATE/TIME OF TEST: 07/05/2016 00:53:03

Diagnostics Check		OK	00:45
Air Blank		0.000	00:45
Control Test		0.075	00:45
Air Blank		0.000	00:46
Subject Sample	# 1	0.175	00:47
Breath Volume		1.222	00:47
Air Blank		0.000	00:48
Air Blank		0.000	00:50
Subject Sample	# 2	0.172	00:51
Breath Volume		1.132	00:51
Air Blank		0.000	00:52
Control Test		0.075	00:52
Air Blank		0.000	00:52
Diagnostics Check		OK	00:53

Data Download Date/Time: 07/31/2016 14:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WIGGLESWORTH,JOE,B
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: PARRISH,R
ARRESTING OFFICER AGENCY: Wildwood PD
LAST AGENCY INSPECTION: 06/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28814080A3
DRY GAS EXPIRY DATE: 11/05/2016
VIOLATION CODE: DUI
ARREST TIME: 19:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:38

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 07/16/2016 21:06:07

Diagnostics Check	OK	20:58
Air Blank	0.000	20:59
Control Test	0.075	20:59
Air Blank	0.000	21:00
Subject Sample # 1	0.113	21:01
Breath Volume	2.109	21:01
Air Blank	0.000	21:02
Air Blank	0.000	21:03
Subject Sample # 2	0.112	21:04
Breath Volume	1.683	21:04
Air Blank	0.000	21:05
Control Test	0.075	21:05
Air Blank	0.000	21:06
Diagnostics Check	OK	21:06

Data Download Date/Time: 07/31/2016 14:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GEOGHEGAN,ASHLEY,M
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: REID,WARREN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28814080A3
DRY GAS EXPIRY DATE: 11/05/2016
VIOLATION CODE: DUI
ARREST TIME: 00:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:34

SUBJECT'S LAST NAME: PAYNE

DATE/TIME OF TEST: 07/23/2016 01:57:49

Diagnostics Check	OK	01:54	
Air Blank	0.000	01:55	
Control Test	0.076	01:55	
Air Blank	0.000	01:56	
Subject Sample	# 1	01:56	Subject Test Refused
Breath Volume	0.000	01:56	
Air Blank	0.000	01:56	
Control Test	0.076	01:57	
Air Blank	0.000	01:57	
Diagnostics Check	OK	01:57	

Data Download Date/Time: 07/31/2016 14:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRANNEN,JUSTIN,C
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: VAUGHN,C
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28814080A3
DRY GAS EXPIRY DATE: 11/05/2016
VIOLATION CODE: DUI
ARREST TIME: 14:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:00

SUBJECT'S LAST NAME: HOWE

DATE/TIME OF TEST: 07/23/2016 15:30:07

Diagnostics Check	OK	15:23	
Air Blank	0.000	15:23	
Control Test	0.077	15:24	
Air Blank	0.000	15:24	
Subject Sample # 1	0.232	15:25	
Breath Volume	2.863	15:25	
Air Blank	0.000	15:25	
Air Blank	0.000	15:27	
Subject Sample # 2	0.230	15:28	
Breath Volume	2.484	15:28	
Air Blank	0.000	15:29	
Control Test	0.074	15:29	Control Outside Tolerance
Air Blank	0.000	15:30	

Data Download Date/Time: 07/31/2016 14:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRANNEN,JUSTIN,C
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: VAUGHN,C
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28814080A3
DRY GAS EXPIRY DATE: 11/05/2016
VIOLATION CODE: DUI
ARREST TIME: 14:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:15

SUBJECT'S LAST NAME: HOWE

DATE/TIME OF TEST: 07/23/2016 15:42:59

Diagnostics Check	OK	15:36
Air Blank	0.000	15:36
Control Test	0.075	15:36
Air Blank	0.000	15:37
Subject Sample # 1	0.238	15:38
Breath Volume	2.550	15:38
Air Blank	0.000	15:39
Air Blank	0.000	15:40
Subject Sample # 2	0.227	15:41
Breath Volume	2.367	15:41
Air Blank	0.000	15:42
Control Test	0.075	15:42
Air Blank	0.000	15:42
Diagnostics Check	OK	15:42

Data Download Date/Time: 07/31/2016 14:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GEOGHEGAN,ASHLEY,M
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: MCCAIG,ASHLEY
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 06/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28814080A3
DRY GAS EXPIRY DATE: 11/05/2016
VIOLATION CODE: DUI
ARREST TIME: 00:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:36

SUBJECT'S LAST NAME: MCCOOK

DATE/TIME OF TEST: 07/29/2016 01:04:48

Diagnostics Check	OK	00:56
Air Blank	0.000	00:57
Control Test	0.075	00:57
Air Blank	0.000	00:58
Subject Sample # 1	0.140	00:58
Breath Volume	1.503	00:58
Air Blank	0.000	00:59
Air Blank	0.000	01:01
Subject Sample # 2	0.136	01:03
Breath Volume	1.656	01:03
Air Blank	0.000	01:03
Control Test	0.076	01:04
Air Blank	0.000	01:04
Diagnostics Check	OK	01:04

Data Download Date/Time: 07/31/2016 14:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCAZZIO,SEAN,J
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: CONRAN,A
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: GILLEENY-SLINGERLAND

DATE/TIME OF TEST: 07/01/2016 23:35:08

Diagnostics Check	OK	23:26
Air Blank	0.000	23:26
Control Test	0.080	23:26
Air Blank	0.000	23:27
Subject Sample # 1	0.141	23:29
Breath Volume	1.460	23:29
Air Blank	0.000	23:30
Air Blank	0.000	23:32
Subject Sample # 2	0.137	23:33
Breath Volume	1.367	23:33
Air Blank	0.000	23:34
Control Test	0.080	23:34
Air Blank	0.000	23:35
Diagnostics Check	OK	23:35

Data Download Date/Time: 07/29/2016 11:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHOETTNER,ADRIAN,J
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: GENTILE,J
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 06/29/2016

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: ALLISON

DATE/TIME OF TEST: 07/02/2016 01:26:07

Diagnostics Check		OK	01:18
Air Blank		0.000	01:18
Control Test		0.079	01:18
Air Blank		0.000	01:19
Subject Sample	# 1	0.182	01:20
Breath Volume		1.359	01:20
Air Blank		0.000	01:21
Air Blank		0.000	01:23
Subject Sample	# 2	0.172	01:24
Breath Volume		2.523	01:24
Air Blank		0.000	01:25
Control Test		0.079	01:25
Air Blank		0.000	01:26
Diagnostics Check		OK	01:26

Data Download Date/Time: 07/29/2016 11:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: METZ,KYLE,J
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: CONRAN,TROOPER
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 13:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:12

SUBJECT'S LAST NAME: VILLANEDA

DATE/TIME OF TEST: 07/02/2016 14:43:08

Diagnostics Check	OK	14:36
Air Blank	0.000	14:37
Control Test	0.079	14:37
Air Blank	0.000	14:37
Subject Sample # 1	0.200	14:38
Breath Volume	1.535	14:38
Air Blank	0.000	14:39
Air Blank	0.000	14:41
Subject Sample # 2	0.200	14:41
Breath Volume	1.847	14:41
Air Blank	0.000	14:42
Control Test	0.078	14:42
Air Blank	0.000	14:43
Diagnostics Check	OK	14:43

Data Download Date/Time: 07/29/2016 11:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCAZZIO,SEAN,J
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: CRANE,R
ARRESTING OFFICER AGENCY: FFWCC (Ocala)
LAST AGENCY INSPECTION: 06/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: BUI
ARREST TIME: 19:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:05

SUBJECT'S LAST NAME: HOWES

DATE/TIME OF TEST: 07/02/2016 20:31:46

Diagnostics Check	OK	20:28	
Air Blank	0.000	20:29	
Control Test	0.080	20:29	
Air Blank	0.000	20:30	
Subject Sample	# 1	20:30	Subject Test Refused
Breath Volume	0.000	20:30	
Air Blank	0.000	20:30	
Control Test	0.080	20:31	
Air Blank	0.000	20:31	
Diagnostics Check	OK	20:31	

Data Download Date/Time: 07/29/2016 11:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCAZZIO,SEAN,J
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: WELLS,C
ARRESTING OFFICER AGENCY: FFWCC (Ocala)
LAST AGENCY INSPECTION: 06/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: BUI
ARREST TIME: 19:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:55

SUBJECT'S LAST NAME: WHITMAN

DATE/TIME OF TEST: 07/02/2016 22:23:26

Diagnostics Check		OK	22:17
Air Blank		0.000	22:17
Control Test		0.079	22:18
Air Blank		0.000	22:18
Subject Sample	# 1	0.246	22:18
Breath Volume		1.707	22:18
Air Blank		0.000	22:19
Air Blank		0.000	22:21
Subject Sample	# 2	0.248	22:21
Breath Volume		1.113	22:21
Air Blank		0.000	22:22
Control Test		0.080	22:22
Air Blank		0.000	22:23
Diagnostics Check		OK	22:23

Data Download Date/Time: 07/29/2016 11:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCAZZIO,SEAN,J
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: THOMASON,D
ARRESTING OFFICER AGENCY: FFWCC (Ocala)
LAST AGENCY INSPECTION: 06/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: BUI
ARREST TIME: 21:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: GUTIERREZ

DATE/TIME OF TEST: 07/02/2016 23:00:23

Diagnostics Check	OK	22:57	
Air Blank	0.000	22:58	
Control Test	0.079	22:58	
Air Blank	0.000	22:58	
Subject Sample	# 1	22:59	Subject Test Refused
Breath Volume	0.000	22:59	
Air Blank	0.000	22:59	
Control Test	0.080	22:59	
Air Blank	0.000	23:00	
Diagnostics Check	OK	23:00	

Data Download Date/Time: 07/29/2016 11:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCAZZIO,SEAN,J
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: CRANE,R
ARRESTING OFFICER AGENCY: FFWCC (Ocala)
LAST AGENCY INSPECTION: 06/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: BUI
ARREST TIME: 17:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:00

SUBJECT'S LAST NAME: SKIDMORE

DATE/TIME OF TEST: 07/03/2016 18:27:04

Diagnostics Check	OK	18:20
Air Blank	0.000	18:21
Control Test	0.080	18:21
Air Blank	0.000	18:22
Subject Sample # 1	0.178	18:22
Breath Volume	2.203	18:22
Air Blank	0.000	18:23
Air Blank	0.000	18:25
Subject Sample # 2	0.178	18:25
Breath Volume	1.851	18:25
Air Blank	0.000	18:26
Control Test	0.080	18:26
Air Blank	0.000	18:26
Diagnostics Check	OK	18:27

Data Download Date/Time: 07/29/2016 11:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCAZZIO,SEAN,J
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: ULRICH,B
ARRESTING OFFICER AGENCY: FFWCC (Ocala)
LAST AGENCY INSPECTION: 06/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: BUI
ARREST TIME: 18:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:25

SUBJECT'S LAST NAME: WINZEL

DATE/TIME OF TEST: 07/03/2016 19:56:37

Diagnostics Check	OK	19:50
Air Blank	0.000	19:50
Control Test	0.080	19:51
Air Blank	0.000	19:51
Subject Sample # 1	0.105	19:52
Breath Volume	2.042	19:52
Air Blank	0.000	19:52
Air Blank	0.000	19:54
Subject Sample # 2	0.111	19:55
Breath Volume	1.925	19:55
Air Blank	0.000	19:55
Control Test	0.079	19:56
Air Blank	0.000	19:56
Diagnostics Check	OK	19:56

Data Download Date/Time: 07/29/2016 11:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCAZZIO,SEAN,J
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: CANFIELD,L
ARRESTING OFFICER AGENCY: FFWCC (Ocala)
LAST AGENCY INSPECTION: 06/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: BUI
ARREST TIME: 18:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:03

SUBJECT'S LAST NAME: STALEY

DATE/TIME OF TEST: 07/03/2016 20:27:05

Diagnostics Check	OK	20:24	
Air Blank	0.000	20:24	
Control Test	0.080	20:25	
Air Blank	0.000	20:25	
Subject Sample	# 1	20:25	Subject Test Refused
Breath Volume	0.000	20:25	
Air Blank	0.000	20:26	
Control Test	0.080	20:26	
Air Blank	0.000	20:27	
Diagnostics Check	OK	20:27	

Data Download Date/Time: 07/29/2016 11:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHOETTNER,ADRIAN,J
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: SUTTLES,J
ARRESTING OFFICER AGENCY: FFWCC (Ocala)
LAST AGENCY INSPECTION: 06/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: BUI
ARREST TIME: 20:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:37

SUBJECT'S LAST NAME: BLAIR

DATE/TIME OF TEST: 07/03/2016 22:13:49

Diagnostics Check	OK	22:10	
Air Blank	0.000	22:11	
Control Test	0.080	22:11	
Air Blank	0.000	22:12	
Subject Sample	# 1	22:12	Subject Test Refused
Breath Volume	0.000	22:12	
Air Blank	0.000	22:12	
Control Test	0.080	22:13	
Air Blank	0.000	22:13	
Diagnostics Check	OK	22:13	

Data Download Date/Time: 07/29/2016 11:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHOETTNER,ADRIAN,J
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: GERENA,G
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 06/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: MONROE

DATE/TIME OF TEST: 07/03/2016 23:11:30

Diagnostics Check	OK	23:05	
Air Blank	0.000	23:05	
Control Test	0.080	23:06	
Air Blank	0.000	23:06	
Subject Sample	# 1 0.098	23:09	Volume Not Met
Breath Volume	0.300	23:09	
Air Blank		23:10	Improper Sample
Air Blank	0.000	23:10	
Control Test	0.080	23:10	
Air Blank	0.000	23:11	
Diagnostics Check	OK	23:11	

Data Download Date/Time: 07/29/2016 11:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIBILIA,CHRISTOPHER,R
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: CRANE,RYAN
ARRESTING OFFICER AGENCY: FFWCC (Ocala)
LAST AGENCY INSPECTION: 06/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: BUI
ARREST TIME: 22:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:01

SUBJECT'S LAST NAME: LAING

DATE/TIME OF TEST: 07/04/2016 23:38:15

Diagnostics Check	OK	23:35	
Air Blank	0.000	23:35	
Control Test	0.080	23:36	
Air Blank	0.000	23:36	
Subject Sample	# 1	23:36	Subject Test Refused
Breath Volume	0.000	23:36	
Air Blank	0.000	23:37	
Control Test	0.080	23:37	
Air Blank	0.000	23:38	
Diagnostics Check	OK	23:38	

Data Download Date/Time: 07/29/2016 11:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TOMLINSON,TERESA,R
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: PANETTI,J
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 06/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 19:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:07

SUBJECT'S LAST NAME: TOMLINSON

DATE/TIME OF TEST: 07/07/2016 20:43:10

Diagnostics Check	OK	20:39	
Air Blank	0.000	20:39	
Control Test	0.079	20:40	
Air Blank	0.000	20:40	
Subject Sample	# 1	20:41	Subject Test Refused
Breath Volume	0.000	20:41	
Air Blank	0.000	20:42	
Control Test	0.080	20:42	
Air Blank	0.000	20:43	
Diagnostics Check	OK	20:43	

Data Download Date/Time: 07/29/2016 11:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RHODE,ERIC,C
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: HORTON,T
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 06/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:28

SUBJECT'S LAST NAME: NOVO

DATE/TIME OF TEST: 07/10/2016 04:17:34

Diagnostics Check	OK	04:10
Air Blank	0.000	04:10
Control Test	0.080	04:10
Air Blank	0.000	04:11
Subject Sample # 1	0.000	04:12
Breath Volume	1.703	04:12
Air Blank	0.000	04:12
Air Blank	0.000	04:14
Subject Sample # 2	0.000	04:16
Breath Volume	1.605	04:16
Air Blank	0.000	04:16
Control Test	0.080	04:16
Air Blank	0.000	04:17
Diagnostics Check	OK	04:17

Data Download Date/Time: 07/29/2016 11:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WING,JOSHUA,M
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: KJEKSTAD,E
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 15:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:22

SUBJECT'S LAST NAME: STODDARD

DATE/TIME OF TEST: 07/13/2016 16:49:38

Diagnostics Check	OK	16:43
Air Blank	0.000	16:43
Control Test	0.080	16:44
Air Blank	0.000	16:44
Subject Sample # 1	0.211	16:45
Breath Volume	1.925	16:45
Air Blank	0.000	16:45
Air Blank	0.000	16:47
Subject Sample # 2	0.213	16:48
Breath Volume	2.074	16:48
Air Blank	0.000	16:48
Control Test	0.080	16:49
Air Blank	0.000	16:49
Diagnostics Check	OK	16:49

Data Download Date/Time: 07/29/2016 11:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOGUE,VINCE,H
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: KJEKSTAD,E
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:36

SUBJECT'S LAST NAME: HART

DATE/TIME OF TEST: 07/14/2016 23:59:24

Diagnostics Check	OK	23:56	
Air Blank	0.000	23:56	
Control Test	0.080	23:57	
Air Blank	0.000	23:57	
Subject Sample	# 1	23:58	Subject Test Refused
Breath Volume	0.000	23:58	
Air Blank	0.000	23:58	
Control Test	0.081	23:58	
Air Blank	0.000	23:59	
Diagnostics Check	OK	23:59	

Data Download Date/Time: 07/29/2016 11:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,CRYSTAL,L
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: ELMORE,P
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:05

SUBJECT'S LAST NAME: KENYON

DATE/TIME OF TEST: 07/15/2016 23:40:12

Diagnostics Check	OK	23:33
Air Blank	0.000	23:33
Control Test	0.080	23:34
Air Blank	0.000	23:34
Subject Sample # 1	0.143	23:35
Breath Volume	3.035	23:35
Air Blank	0.000	23:36
Air Blank	0.000	23:38
Subject Sample # 2	0.144	23:38
Breath Volume	3.371	23:38
Air Blank	0.000	23:39
Control Test	0.080	23:39
Air Blank	0.000	23:40
Diagnostics Check	OK	23:40

Data Download Date/Time: 07/29/2016 11:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOGUE,VINCE,H
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: FELIX,R
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 06/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:14

SUBJECT'S LAST NAME: ELLIOTT

DATE/TIME OF TEST: 07/22/2016 21:42:39

Diagnostics Check	OK	21:35
Air Blank	0.000	21:35
Control Test	0.080	21:35
Air Blank	0.000	21:36
Subject Sample # 1	0.154	21:37
Breath Volume	2.003	21:37
Air Blank	0.000	21:38
Air Blank	0.000	21:40
Subject Sample # 2	0.157	21:41
Breath Volume	2.644	21:41
Air Blank	0.000	21:41
Control Test	0.079	21:42
Air Blank	0.000	21:42
Diagnostics Check	OK	21:42

Data Download Date/Time: 07/29/2016 11:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCAZZIO,SEAN,J
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: WILLIAMS,K
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 07/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: MARKS

DATE/TIME OF TEST: 07/31/2016 00:53:47

Diagnostics Check	OK	00:50	
Air Blank	0.000	00:51	
Control Test	0.080	00:51	
Air Blank	0.000	00:52	
Subject Sample	# 1	00:52	Subject Test Refused
Breath Volume	0.000	00:52	
Air Blank	0.000	00:52	
Control Test	0.080	00:53	
Air Blank	0.000	00:53	
Diagnostics Check	OK	00:53	

Data Download Date/Time: 08/22/2016 09:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ, DAVID,
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: GONZALEZ, DAVID
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 06/27/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 07:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:00

SUBJECT'S LAST NAME: BELL

DATE/TIME OF TEST: 07/07/2016 08:12:55

Diagnostics Check	OK	08:06
Air Blank	0.000	08:06
Control Test	0.081	08:07
Air Blank	0.000	08:07
Subject Sample # 1	0.153	08:08
Breath Volume	2.789	08:08
Air Blank	0.000	08:09
Air Blank	0.000	08:10
Subject Sample # 2	0.150	08:11
Breath Volume	2.613	08:11
Air Blank	0.000	08:12
Control Test	0.081	08:12
Air Blank	0.000	08:12
Diagnostics Check	OK	08:12

Data Download Date/Time: 07/27/2016 07:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COURSEY,SARAH,M
BREATH TEST OPERATOR AGENCY: Tavares PD
ARRESTING OFFICER: JOHNSON,CODY
ARRESTING OFFICER AGENCY: Tavares PD
LAST AGENCY INSPECTION: 06/27/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 16:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:04

SUBJECT'S LAST NAME: DENNIS

DATE/TIME OF TEST: 07/14/2016 17:44:37

Diagnostics Check	OK	17:38
Air Blank	0.000	17:38
Control Test	0.081	17:38
Air Blank	0.000	17:39
Subject Sample # 1	0.201	17:40
Breath Volume	2.914	17:40
Air Blank	0.000	17:40
Air Blank	0.000	17:42
Subject Sample # 2	0.189	17:43
Breath Volume	2.121	17:43
Air Blank	0.000	17:43
Control Test	0.080	17:44
Air Blank	0.000	17:44
Diagnostics Check	OK	17:44

Data Download Date/Time: 07/27/2016 07:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARRETT,C,
BREATH TEST OPERATOR AGENCY: Clermont PD
ARRESTING OFFICER: WEGNER,B
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 07/27/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 17:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:40

SUBJECT'S LAST NAME: FENNER

DATE/TIME OF TEST: 07/30/2016 18:10:11

Diagnostics Check	OK	18:07	
Air Blank	0.000	18:07	
Control Test	0.080	18:08	
Air Blank	0.000	18:08	
Subject Sample	# 1	18:08	Subject Test Refused
Breath Volume	0.000	18:08	
Air Blank	0.000	18:09	
Control Test	0.080	18:09	
Air Blank	0.000	18:10	
Diagnostics Check	OK	18:10	

Data Download Date/Time: 08/29/2016 10:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,ETHAN,
BREATH TEST OPERATOR AGENCY: Clermont PD
ARRESTING OFFICER: BELL,ETHAN
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 07/27/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:17

SUBJECT'S LAST NAME: BAIRD

DATE/TIME OF TEST: 07/31/2016 01:44:57

Diagnostics Check		OK	01:38
Air Blank		0.000	01:38
Control Test		0.081	01:38
Air Blank		0.000	01:39
Subject Sample	# 1	0.241	01:40
Breath Volume		2.753	01:40
Air Blank		0.000	01:40
Air Blank		0.000	01:42
Subject Sample	# 2	0.244	01:43
Breath Volume		3.609	01:43
Air Blank		0.000	01:44
Control Test		0.082	01:44
Air Blank		0.000	01:44
Diagnostics Check		OK	01:44

Data Download Date/Time: 08/29/2016 10:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRECO,CHARLES,R
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: HART,ROBERT
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 06/27/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1766638
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: LANE

DATE/TIME OF TEST: 07/04/2016 22:10:49

Diagnostics Check	OK	22:04
Air Blank	0.000	22:04
Control Test	0.079	22:04
Air Blank	0.000	22:05
Subject Sample # 1	0.043	22:06
Breath Volume	1.691	22:06
Air Blank	0.000	22:06
Air Blank	0.000	22:08
Subject Sample # 2	0.042	22:09
Breath Volume	1.847	22:09
Air Blank	0.000	22:09
Control Test	0.079	22:10
Air Blank	0.000	22:10
Diagnostics Check	OK	22:10

Data Download Date/Time: 07/27/2016 10:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,ETHAN,
BREATH TEST OPERATOR AGENCY: Clermont PD
ARRESTING OFFICER: BELL,ETHAN
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 06/27/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1766638
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: FRIEKO

DATE/TIME OF TEST: 07/07/2016 00:34:55

Diagnostics Check	OK	00:31	
Air Blank	0.000	00:31	
Control Test	0.078	00:31	
Air Blank	0.000	00:32	
Subject Sample	# 1	00:33	Subject Test Refused
Breath Volume	0.000	00:33	
Air Blank	0.000	00:34	
Control Test	0.079	00:34	
Air Blank	0.000	00:34	
Diagnostics Check	OK	00:34	

Data Download Date/Time: 07/27/2016 10:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OLDDHAM,TAYLOR,D
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: CORMAN,GERALD
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 06/27/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1766638
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 07/13/2016 02:13:01

Diagnostics Check	OK	02:06
Air Blank	0.000	02:06
Control Test	0.078	02:07
Air Blank	0.000	02:07
Subject Sample # 1	0.163	02:08
Breath Volume	2.343	02:08
Air Blank	0.000	02:09
Air Blank	0.000	02:10
Subject Sample # 2	0.156	02:11
Breath Volume	3.023	02:11
Air Blank	0.000	02:12
Control Test	0.077	02:12
Air Blank	0.000	02:12
Diagnostics Check	OK	02:13

Data Download Date/Time: 07/27/2016 10:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLOOM,NICHOLAS,A
BREATH TEST OPERATOR AGENCY: Clermont PD
ARRESTING OFFICER: BLOOM,NICHOLAS
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 06/27/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1766638
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:12

SUBJECT'S LAST NAME: GORSLINE

DATE/TIME OF TEST: 07/15/2016 01:39:29

Diagnostics Check		OK	01:32
Air Blank		0.000	01:33
Control Test		0.079	01:33
Air Blank		0.000	01:33
Subject Sample	# 1	0.180	01:34
Breath Volume		2.535	01:34
Air Blank		0.000	01:35
Air Blank		0.000	01:37
Subject Sample	# 2	0.171	01:37
Breath Volume		1.699	01:37
Air Blank		0.000	01:38
Control Test		0.078	01:38
Air Blank		0.000	01:39
Diagnostics Check		OK	01:39

Data Download Date/Time: 07/27/2016 10:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLOOM,NICHOLAS,A
BREATH TEST OPERATOR AGENCY: Clermont PD
ARRESTING OFFICER: BLOOM,NICHOLAS
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 06/27/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1766638
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: FULLER

DATE/TIME OF TEST: 07/17/2016 03:42:11

Diagnostics Check	OK	03:35
Air Blank	0.000	03:36
Control Test	0.079	03:36
Air Blank	0.000	03:37
Subject Sample # 1	0.000	03:37
Breath Volume	1.550	03:37
Air Blank	0.000	03:38
Air Blank	0.000	03:40
Subject Sample # 2	0.000	03:40
Breath Volume	1.851	03:40
Air Blank	0.000	03:41
Control Test	0.079	03:41
Air Blank	0.000	03:42
Diagnostics Check	OK	03:42

Data Download Date/Time: 07/27/2016 10:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,CHRIS,L
BREATH TEST OPERATOR AGENCY: Clermont PD
ARRESTING OFFICER: AVILA,JEREMIAH
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 06/27/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1766638
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:22

SUBJECT'S LAST NAME: PATRICK

DATE/TIME OF TEST: 07/22/2016 01:50:08

Diagnostics Check	OK	01:42
Air Blank	0.000	01:43
Control Test	0.079	01:43
Air Blank	0.000	01:43
Subject Sample # 1	0.193	01:45
Breath Volume	2.394	01:45
Air Blank	0.000	01:45
Air Blank	0.000	01:47
Subject Sample # 2	0.201	01:48
Breath Volume	3.062	01:48
Air Blank	0.000	01:49
Control Test	0.078	01:49
Air Blank	0.000	01:50
Diagnostics Check	OK	01:50

Data Download Date/Time: 07/27/2016 10:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEVENS, JEREMY, L
BREATH TEST OPERATOR AGENCY: Leesburg PD
ARRESTING OFFICER: WHITLEY, TRAVIS
ARRESTING OFFICER AGENCY: Leesburg PD
LAST AGENCY INSPECTION: 06/01/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A1
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 18:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:20

SUBJECT'S LAST NAME: HARRISON

DATE/TIME OF TEST: 07/16/2016 19:43:10

Diagnostics Check	OK	19:40	
Air Blank	0.000	19:40	
Control Test	0.079	19:41	
Air Blank	0.000	19:41	
Subject Sample	# 1	19:41	Subject Test Refused
Breath Volume	0.000	19:41	
Air Blank	0.000	19:42	
Control Test	0.079	19:42	
Air Blank	0.000	19:43	
Diagnostics Check	OK	19:43	

Data Download Date/Time: 07/25/2016 02:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000825

INSTRUMENT REGISTERED TO: MOUNT DORA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAYLOR,KEITH,A
BREATH TEST OPERATOR AGENCY: Mount Dora PD
ARRESTING OFFICER: SMITH,ANGELA
ARRESTING OFFICER AGENCY: Mount Dora PD
LAST AGENCY INSPECTION: 06/27/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1841081
DRY GAS EXPIRY DATE: 04/08/2018
VIOLATION CODE: DUI
ARREST TIME: 01:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:04

SUBJECT'S LAST NAME: DE LA CRUZ

DATE/TIME OF TEST: 07/02/2016 02:57:56

Diagnostics Check	OK	02:54	
Air Blank	0.000	02:55	
Control Test	0.079	02:55	
Air Blank	0.000	02:56	
Subject Sample	# 1	02:56	Subject Test Refused
Breath Volume	0.000	02:56	
Air Blank	0.000	02:57	
Control Test	0.079	02:57	
Air Blank	0.000	02:57	
Diagnostics Check	OK	02:57	

Data Download Date/Time: 07/31/2016 23:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000825

INSTRUMENT REGISTERED TO: MOUNT DORA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,ANGELA,R
BREATH TEST OPERATOR AGENCY: Mount Dora PD
ARRESTING OFFICER: ANDERSON,N
ARRESTING OFFICER AGENCY: Mount Dora PD
LAST AGENCY INSPECTION: 06/27/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1841081
DRY GAS EXPIRY DATE: 04/08/2018
VIOLATION CODE: DUI
ARREST TIME: 21:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:34

SUBJECT'S LAST NAME: CUSIMANO

DATE/TIME OF TEST: 07/02/2016 22:08:08

Diagnostics Check		OK	21:54	
Air Blank		0.000	21:55	
Control Test		0.078	21:55	
Air Blank		0.000	21:55	
Subject Sample	# 1	0.211	21:59	Volume Not Met
Breath Volume		0.308	21:59	
Air Blank		0.000	21:59	
Air Blank		0.000	22:01	
Subject Sample	# 2	0.203	22:03	
Breath Volume		2.265	22:03	
Air Blank		0.000	22:04	
Air Blank		0.000	22:06	
Subject Sample	# 3	0.187	22:06	
Breath Volume		2.007	22:06	
Air Blank		0.000	22:07	
Control Test		0.076	22:07	
Air Blank		0.000	22:08	
Diagnostics Check		OK	22:08	

Data Download Date/Time: 07/31/2016 23:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000825

INSTRUMENT REGISTERED TO: MOUNT DORA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FIGUEROA,H,
BREATH TEST OPERATOR AGENCY: Mount Dora PD
ARRESTING OFFICER: FIGUEROA,H
ARRESTING OFFICER AGENCY: Mount Dora PD
LAST AGENCY INSPECTION: 06/27/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1841081
DRY GAS EXPIRY DATE: 04/08/2018
VIOLATION CODE: DUI
ARREST TIME: 06:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:18

SUBJECT'S LAST NAME: BRONSON

DATE/TIME OF TEST: 07/05/2016 06:49:50

Diagnostics Check	OK	06:42
Air Blank	0.000	06:43
Control Test	0.078	06:43
Air Blank	0.000	06:43
Subject Sample # 1	0.000	06:45
Breath Volume	2.304	06:45
Air Blank	0.000	06:45
Air Blank	0.000	06:47
Subject Sample # 2	0.000	06:48
Breath Volume	1.234	06:48
Air Blank	0.000	06:48
Control Test	0.076	06:49
Air Blank	0.000	06:49
Diagnostics Check	OK	06:49

Data Download Date/Time: 07/31/2016 23:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000825

INSTRUMENT REGISTERED TO: MOUNT DORA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAYLOR,KEITH,A
BREATH TEST OPERATOR AGENCY: Mount Dora PD
ARRESTING OFFICER: VAUGHN,MICHAEL
ARRESTING OFFICER AGENCY: Mount Dora PD
LAST AGENCY INSPECTION: 06/27/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1841081
DRY GAS EXPIRY DATE: 04/08/2018
VIOLATION CODE: DUI
ARREST TIME: 22:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:18

SUBJECT'S LAST NAME: SPILLIARD

DATE/TIME OF TEST: 07/15/2016 22:47:52

Diagnostics Check	OK	22:41
Air Blank	0.000	22:41
Control Test	0.078	22:42
Air Blank	0.000	22:42
Subject Sample # 1	0.253	22:43
Breath Volume	2.812	22:43
Air Blank	0.000	22:44
Air Blank	0.000	22:45
Subject Sample # 2	0.244	22:46
Breath Volume	2.464	22:46
Air Blank	0.000	22:46
Control Test	0.077	22:47
Air Blank	0.000	22:47
Diagnostics Check	OK	22:47

Data Download Date/Time: 07/31/2016 23:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000825

INSTRUMENT REGISTERED TO: MOUNT DORA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAYLOR,KEITH,A
BREATH TEST OPERATOR AGENCY: Mount Dora PD
ARRESTING OFFICER: TAYLOR,KEITH
ARRESTING OFFICER AGENCY: Mount Dora PD
LAST AGENCY INSPECTION: 06/27/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1841081
DRY GAS EXPIRY DATE: 04/08/2018
VIOLATION CODE: DUI
ARREST TIME: 02:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: RYDBERG

DATE/TIME OF TEST: 07/17/2016 03:09:39

Diagnostics Check	OK	03:06	
Air Blank	0.000	03:07	
Control Test	0.078	03:07	
Air Blank	0.000	03:07	
Subject Sample	# 1	03:08	Subject Test Refused
Breath Volume	0.000	03:08	
Air Blank	0.000	03:08	
Control Test	0.077	03:09	
Air Blank	0.000	03:09	
Diagnostics Check	OK	03:09	

Data Download Date/Time: 07/31/2016 23:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000825

INSTRUMENT REGISTERED TO: MOUNT DORA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAYLOR,KEITH,A
BREATH TEST OPERATOR AGENCY: Mount Dora PD
ARRESTING OFFICER: TAYLOR,KEITH
ARRESTING OFFICER AGENCY: Mount Dora PD
LAST AGENCY INSPECTION: 06/27/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1841081
DRY GAS EXPIRY DATE: 04/08/2018
VIOLATION CODE: DUI
ARREST TIME: 02:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: SEGURA

DATE/TIME OF TEST: 07/27/2016 03:31:58

Diagnostics Check	OK	03:25
Air Blank	0.000	03:26
Control Test	0.078	03:26
Air Blank	0.000	03:26
Subject Sample # 1	0.113	03:27
Breath Volume	1.738	03:27
Air Blank	0.000	03:28
Air Blank	0.000	03:29
Subject Sample # 2	0.109	03:30
Breath Volume	1.386	03:30
Air Blank	0.000	03:31
Control Test	0.079	03:31
Air Blank	0.000	03:31
Diagnostics Check	OK	03:31

Data Download Date/Time: 07/31/2016 23:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000825

INSTRUMENT REGISTERED TO: MOUNT DORA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,ANGELA,R
BREATH TEST OPERATOR AGENCY: Mount Dora PD
ARRESTING OFFICER: VAUGHN,M
ARRESTING OFFICER AGENCY: Mount Dora PD
LAST AGENCY INSPECTION: 06/27/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1841081
DRY GAS EXPIRY DATE: 04/08/2018
VIOLATION CODE: DUI
ARREST TIME: 20:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:03

SUBJECT'S LAST NAME: BRADLEY

DATE/TIME OF TEST: 07/30/2016 21:30:24

Diagnostics Check	OK	21:23
Air Blank	0.000	21:24
Control Test	0.078	21:24
Air Blank	0.000	21:25
Subject Sample # 1	0.179	21:25
Breath Volume	2.753	21:25
Air Blank	0.000	21:26
Air Blank	0.000	21:28
Subject Sample # 2	0.177	21:28
Breath Volume	1.929	21:28
Air Blank	0.000	21:29
Control Test	0.077	21:29
Air Blank	0.000	21:30
Diagnostics Check	OK	21:30

Data Download Date/Time: 07/31/2016 23:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000825

INSTRUMENT REGISTERED TO: MOUNT DORA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,ANGELA,R
BREATH TEST OPERATOR AGENCY: Mount Dora PD
ARRESTING OFFICER: TAYLOR,K
ARRESTING OFFICER AGENCY: Mount Dora PD
LAST AGENCY INSPECTION: 06/27/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1841081
DRY GAS EXPIRY DATE: 04/08/2018
VIOLATION CODE: DUI
ARREST TIME: 03:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: SCHUTTE

DATE/TIME OF TEST: 07/31/2016 03:58:27

Diagnostics Check	OK	03:46	
Air Blank	0.000	03:46	
Control Test	0.078	03:47	
Air Blank	0.000	03:47	
Subject Sample	# 1	03:50	No Sample Provided
Breath Volume	0.000	03:50	
Air Blank	0.000	03:51	
Air Blank	0.000	03:53	
Subject Sample	# 2	0.151	03:53
Breath Volume	2.160	03:53	
Air Blank	0.000	03:54	
Air Blank	0.000	03:56	
Subject Sample	# 3	0.151	03:56
Breath Volume	2.257	03:56	
Air Blank	0.000	03:57	
Control Test	0.078	03:57	
Air Blank	0.000	03:58	
Diagnostics Check	OK	03:58	

Data Download Date/Time: 07/31/2016 23:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000825

INSTRUMENT REGISTERED TO: MOUNT DORA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,ANGELA,R
BREATH TEST OPERATOR AGENCY: Mount Dora PD
ARRESTING OFFICER: ANDERSON,NEAL
ARRESTING OFFICER AGENCY: Mount Dora PD
LAST AGENCY INSPECTION: 06/27/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1841081
DRY GAS EXPIRY DATE: 04/08/2018
VIOLATION CODE: DUI
ARREST TIME: 03:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:36

SUBJECT'S LAST NAME: PALACIOS

DATE/TIME OF TEST: 07/31/2016 04:19:46

Diagnostics Check	OK	04:13
Air Blank	0.000	04:13
Control Test	0.078	04:14
Air Blank	0.000	04:14
Subject Sample # 1	0.103	04:15
Breath Volume	3.343	04:15
Air Blank	0.000	04:15
Air Blank	0.000	04:17
Subject Sample # 2	0.101	04:18
Breath Volume	3.574	04:18
Air Blank	0.000	04:18
Control Test	0.078	04:19
Air Blank	0.000	04:19
Diagnostics Check	OK	04:19

Data Download Date/Time: 07/31/2016 23:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONGELUZZO, MICHAEL, N
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: ROBERTS, SGT
ARRESTING OFFICER AGENCY: Dunnellon PD
LAST AGENCY INSPECTION: 07/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 83550
DRY GAS EXPIRY DATE: 09/26/2018
VIOLATION CODE: DUI
ARREST TIME: 13:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:25

SUBJECT'S LAST NAME: HANDLER

DATE/TIME OF TEST: 07/22/2016 14:52:19

Diagnostics Check	OK	14:45
Air Blank	0.000	14:46
Control Test	0.079	14:46
Air Blank	0.000	14:46
Subject Sample # 1	0.303	14:47
Breath Volume	2.082	14:47
Air Blank	0.000	14:48
Air Blank	0.000	14:50
Subject Sample # 2	0.299	14:50
Breath Volume	1.914	14:50
Air Blank	0.000	14:51
Control Test	0.078	14:51
Air Blank	0.000	14:52
Diagnostics Check	OK	14:52

Data Download Date/Time: 08/26/2016 12:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GAGO-
CHUQUILLANQUI,ALBERTO,E
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: STRUBLE,STEVEN
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 07/21/2016

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 83550
DRY GAS EXPIRY DATE: 09/26/2018
VIOLATION CODE: DUI
ARREST TIME: 01:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:39

SUBJECT'S LAST NAME: LANGFORD

DATE/TIME OF TEST: 07/23/2016 02:05:20

Diagnostics Check	OK	01:59
Air Blank	0.000	01:59
Control Test	0.079	01:59
Air Blank	0.000	02:00
Subject Sample # 1	0.189	02:00
Breath Volume	1.523	02:00
Air Blank	0.000	02:01
Air Blank	0.000	02:03
Subject Sample # 2	0.189	02:03
Breath Volume	2.320	02:03
Air Blank	0.000	02:04
Control Test	0.078	02:04
Air Blank	0.000	02:05
Diagnostics Check	OK	02:05

Data Download Date/Time: 08/26/2016 12:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,GARY,2
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: MILLER,GARY
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 07/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 83550
DRY GAS EXPIRY DATE: 09/26/2018
VIOLATION CODE: DUI
ARREST TIME: 02:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:04

SUBJECT'S LAST NAME: MOLINA

DATE/TIME OF TEST: 07/23/2016 03:30:51

Diagnostics Check	OK	03:24
Air Blank	0.000	03:25
Control Test	0.079	03:25
Air Blank	0.000	03:25
Subject Sample # 1	0.164	03:26
Breath Volume	2.414	03:26
Air Blank	0.000	03:27
Air Blank	0.000	03:28
Subject Sample # 2	0.171	03:29
Breath Volume	1.859	03:29
Air Blank	0.000	03:29
Control Test	0.079	03:30
Air Blank	0.000	03:30
Diagnostics Check	OK	03:30

Data Download Date/Time: 08/26/2016 12:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,GARY,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: EDWARDS,ALAN
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 07/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 83550
DRY GAS EXPIRY DATE: 09/26/2018
VIOLATION CODE: DUI
ARREST TIME: 02:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: WARD

DATE/TIME OF TEST: 07/23/2016 04:12:16

Diagnostics Check	OK	04:05
Air Blank	0.000	04:06
Control Test	0.079	04:06
Air Blank	0.000	04:07
Subject Sample # 1	0.195	04:07
Breath Volume	1.890	04:07
Air Blank	0.000	04:08
Air Blank	0.000	04:10
Subject Sample # 2	0.196	04:10
Breath Volume	1.597	04:10
Air Blank	0.000	04:11
Control Test	0.079	04:11
Air Blank	0.000	04:12
Diagnostics Check	OK	04:12

Data Download Date/Time: 08/26/2016 12:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITE,CHRISTOPHER,C
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: LAROCQUE,ANNEMARIE
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 07/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 83550
DRY GAS EXPIRY DATE: 09/26/2018
VIOLATION CODE: DUI
ARREST TIME: 02:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:07

SUBJECT'S LAST NAME: MARQUEZ RAMIREZ

DATE/TIME OF TEST: 07/25/2016 03:46:26

Diagnostics Check	OK	03:39
Air Blank	0.000	03:39
Control Test	0.080	03:39
Air Blank	0.000	03:40
Subject Sample # 1	0.178	03:41
Breath Volume	3.429	03:41
Air Blank	0.000	03:42
Air Blank	0.000	03:44
Subject Sample # 2	0.170	03:44
Breath Volume	2.421	03:44
Air Blank	0.000	03:45
Control Test	0.078	03:45
Air Blank	0.000	03:46
Diagnostics Check	OK	03:46

Data Download Date/Time: 08/26/2016 12:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,GARY,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: MILLER,GARY
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 07/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 83550
DRY GAS EXPIRY DATE: 09/26/2018
VIOLATION CODE: DUI
ARREST TIME: 19:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:57

SUBJECT'S LAST NAME: KUSION

DATE/TIME OF TEST: 07/25/2016 20:25:12

Diagnostics Check	OK	20:18
Air Blank	0.000	20:19
Control Test	0.079	20:19
Air Blank	0.000	20:19
Subject Sample # 1	0.238	20:20
Breath Volume	2.140	20:20
Air Blank	0.000	20:21
Air Blank	0.000	20:23
Subject Sample # 2	0.240	20:23
Breath Volume	2.492	20:23
Air Blank	0.000	20:24
Control Test	0.079	20:24
Air Blank	0.000	20:25
Diagnostics Check	OK	20:25

Data Download Date/Time: 08/26/2016 12:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEWART,RODNEY,L
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: SIMS,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 07/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 83550
DRY GAS EXPIRY DATE: 09/26/2018
VIOLATION CODE: DUI
ARREST TIME: 18:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:20

SUBJECT'S LAST NAME: STEWART

DATE/TIME OF TEST: 07/26/2016 19:47:12

Diagnostics Check	OK	19:40
Air Blank	0.000	19:41
Control Test	0.080	19:41
Air Blank	0.000	19:42
Subject Sample # 1	0.080	19:42
Breath Volume	1.824	19:42
Air Blank	0.000	19:43
Air Blank	0.000	19:45
Subject Sample # 2	0.077	19:45
Breath Volume	1.789	19:45
Air Blank	0.000	19:46
Control Test	0.080	19:46
Air Blank	0.000	19:47
Diagnostics Check	OK	19:47

Data Download Date/Time: 08/26/2016 12:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITMORE,JAMES,D
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: WHITMORE,JAMES
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 07/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1766638
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: MATHEWSON

DATE/TIME OF TEST: 07/29/2016 02:30:25

Diagnostics Check	OK	02:23
Air Blank	0.000	02:23
Control Test	0.081	02:24
Air Blank	0.000	02:24
Subject Sample # 1	0.157	02:25
Breath Volume	1.859	02:25
Air Blank	0.000	02:26
Air Blank	0.000	02:27
Subject Sample # 2	0.166	02:28
Breath Volume	2.542	02:28
Air Blank	0.000	02:29
Control Test	0.080	02:29
Air Blank	0.000	02:30
Diagnostics Check	OK	02:30

Data Download Date/Time: 08/26/2016 12:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARK, MATTHEW, J
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: PARK, MATTHEW
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 07/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 83550
DRY GAS EXPIRY DATE: 09/26/2018
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:54

SUBJECT'S LAST NAME: WALDEN

DATE/TIME OF TEST: 07/29/2016 04:26:46

Diagnostics Check	OK	04:18
Air Blank	0.000	04:19
Control Test	0.080	04:19
Air Blank	0.000	04:20
Subject Sample # 1	0.202	04:22
Breath Volume	3.148	04:22
Air Blank	0.000	04:22
Air Blank	0.000	04:24
Subject Sample # 2	0.195	04:25
Breath Volume	2.789	04:25
Air Blank	0.000	04:25
Control Test	0.077	04:26
Air Blank	0.000	04:26
Diagnostics Check	OK	04:26

Data Download Date/Time: 08/26/2016 12:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALLER,JESSICA,G
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: GALLER,JESSICA
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 07/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 83550
DRY GAS EXPIRY DATE: 09/26/2018
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:37

SUBJECT'S LAST NAME: BIKOWITZ

DATE/TIME OF TEST: 07/30/2016 03:09:15

Diagnostics Check		OK	02:57	
Air Blank		0.000	02:58	
Control Test		0.080	02:58	
Air Blank		0.000	02:58	
Subject Sample	# 1	0.110	03:02	Volume Not Met
Breath Volume		0.074	03:02	
Air Blank		0.000	03:02	
Air Blank		0.000	03:04	
Subject Sample	# 2	0.127	03:07	Volume Not Met
Breath Volume		0.261	03:07	
Air Blank		0.000	03:08	
Control Test		0.078	03:08	
Air Blank		0.000	03:09	
Diagnostics Check		OK	03:09	

Data Download Date/Time: 08/26/2016 12:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HULL,ROBERT,E
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: HULL,ROBERT
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 07/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1766638
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 10:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: BEVIL

DATE/TIME OF TEST: 07/31/2016 00:36:38

Diagnostics Check		OK	00:30
Air Blank		0.000	00:30
Control Test		0.081	00:31
Air Blank		0.000	00:31
Subject Sample	# 1	0.000	00:32
Breath Volume		2.000	00:32
Air Blank		0.000	00:32
Air Blank		0.000	00:34
Subject Sample	# 2	0.000	00:35
Breath Volume		3.144	00:35
Air Blank		0.000	00:35
Control Test		0.078	00:36
Air Blank		0.000	00:36
Diagnostics Check		OK	00:36

Data Download Date/Time: 08/26/2016 12:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTON,TIMOTHY,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: MARTON,TIMOTHY
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 07/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1766638
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:18

SUBJECT'S LAST NAME: AMMEND

DATE/TIME OF TEST: 07/31/2016 01:01:01

Diagnostics Check	OK	00:54
Air Blank	0.000	00:55
Control Test	0.080	00:55
Air Blank	0.000	00:56
Subject Sample # 1	0.285	00:56
Breath Volume	1.714	00:56
Air Blank	0.000	00:57
Air Blank	0.000	00:58
Subject Sample # 2	0.300	00:59
Breath Volume	1.550	00:59
Air Blank	0.000	01:00
Control Test	0.078	01:00
Air Blank	0.000	01:00
Diagnostics Check	OK	01:01

Data Download Date/Time: 08/26/2016 12:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LETSON,REBECCA,G
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: LETSON,REBECCA
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 07/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 83550
DRY GAS EXPIRY DATE: 09/26/2018
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:01

SUBJECT'S LAST NAME: KRAUSE JR

DATE/TIME OF TEST: 07/31/2016 02:34:14

Diagnostics Check	OK	02:27
Air Blank	0.000	02:28
Control Test	0.080	02:28
Air Blank	0.000	02:29
Subject Sample # 1	0.202	02:29
Breath Volume	1.347	02:29
Air Blank	0.000	02:30
Air Blank	0.000	02:32
Subject Sample # 2	0.196	02:32
Breath Volume	1.566	02:32
Air Blank	0.000	02:33
Control Test	0.079	02:33
Air Blank	0.000	02:34
Diagnostics Check	OK	02:34

Data Download Date/Time: 08/26/2016 12:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNYDER,MELONIE,E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SERRA-RODRIGUEZ,M
ARRESTING OFFICER AGENCY: Macdill Air Force Ba
LAST AGENCY INSPECTION: 07/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 07:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:28

SUBJECT'S LAST NAME: COX

DATE/TIME OF TEST: 07/13/2016 08:56:02

Diagnostics Check	OK	08:49
Air Blank	0.000	08:49
Control Test	0.078	08:50
Air Blank	0.000	08:50
Subject Sample # 1	0.236	08:51
Breath Volume	3.324	08:51
Air Blank	0.000	08:51
Air Blank	0.000	08:53
Subject Sample # 2	0.231	08:54
Breath Volume	3.523	08:54
Air Blank	0.000	08:55
Control Test	0.078	08:55
Air Blank	0.000	08:55
Diagnostics Check	OK	08:56

Data Download Date/Time: 07/13/2016 14:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: NIEUWENDAM,R
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:57

SUBJECT'S LAST NAME: CANADY

DATE/TIME OF TEST: 07/14/2016 01:28:50

Diagnostics Check		OK	01:17	
Air Blank		0.000	01:17	
Control Test		0.079	01:18	
Air Blank		0.000	01:18	
Subject Sample	# 1	0.086	01:21	Volume Not Met
Breath Volume		0.492	01:21	
Air Blank		0.000	01:22	
Air Blank		0.000	01:24	
Subject Sample	# 2	0.085	01:27	Volume Not Met
Breath Volume		0.425	01:27	
Air Blank		0.000	01:27	
Control Test		0.078	01:28	
Air Blank		0.000	01:28	
Diagnostics Check		OK	01:28	

Data Download Date/Time: 08/10/2016 20:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: NIEUWENDAM,R
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:57

SUBJECT'S LAST NAME: CANADY

DATE/TIME OF TEST: 07/14/2016 01:44:25

Diagnostics Check		OK	01:32	
Air Blank		0.000	01:33	
Control Test		0.078	01:33	
Air Blank		0.000	01:34	
Subject Sample	# 1	0.091	01:37	Volume Not Met
Breath Volume		1.007	01:37	
Air Blank		0.000	01:37	
Air Blank		0.000	01:39	
Subject Sample	# 2	0.086	01:42	Volume Not Met
Breath Volume		0.988	01:42	
Air Blank		0.000	01:43	
Control Test		0.078	01:43	
Air Blank		0.000	01:44	
Diagnostics Check		OK	01:44	

Data Download Date/Time: 08/10/2016 20:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: NIEUWENDAM,R
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:57

SUBJECT'S LAST NAME: CANADY

DATE/TIME OF TEST: 07/14/2016 02:03:20

Diagnostics Check		OK	01:51	
Air Blank		0.000	01:52	
Control Test		0.077	01:52	
Air Blank		0.000	01:53	
Subject Sample	# 1	0.085	01:56	Volume Not Met
Breath Volume		1.066	01:56	
Air Blank		0.000	01:56	
Air Blank		0.000	01:58	
Subject Sample	# 2	0.080	02:01	Volume Not Met
Breath Volume		0.015	02:01	
Air Blank		0.000	02:02	
Control Test		0.078	02:02	
Air Blank		0.000	02:03	
Diagnostics Check		OK	02:03	

Data Download Date/Time: 08/10/2016 20:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DE JESUS,V
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:17

SUBJECT'S LAST NAME: PEREZ-RODRIGUEZ

DATE/TIME OF TEST: 07/14/2016 04:44:18

Diagnostics Check	OK	04:37
Air Blank	0.000	04:38
Control Test	0.078	04:38
Air Blank	0.000	04:38
Subject Sample # 1	0.195	04:39
Breath Volume	1.292	04:39
Air Blank	0.000	04:39
Air Blank	0.000	04:41
Subject Sample # 2	0.205	04:42
Breath Volume	1.851	04:42
Air Blank	0.000	04:43
Control Test	0.077	04:43
Air Blank	0.000	04:44
Diagnostics Check	OK	04:44

Data Download Date/Time: 08/10/2016 20:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:34

SUBJECT'S LAST NAME: BORRECA

DATE/TIME OF TEST: 07/15/2016 02:01:30

Diagnostics Check	OK	01:54
Air Blank	0.000	01:55
Control Test	0.078	01:55
Air Blank	0.000	01:56
Subject Sample # 1	0.232	01:57
Breath Volume	2.500	01:57
Air Blank	0.000	01:57
Air Blank	0.000	01:59
Subject Sample # 2	0.220	02:00
Breath Volume	2.167	02:00
Air Blank	0.000	02:00
Control Test	0.077	02:00
Air Blank	0.000	02:01
Diagnostics Check	OK	02:01

Data Download Date/Time: 08/10/2016 20:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY,PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GLEATON,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:25

SUBJECT'S LAST NAME: FLINT

DATE/TIME OF TEST: 07/15/2016 04:54:31

Diagnostics Check	OK	04:48
Air Blank	0.000	04:48
Control Test	0.078	04:49
Air Blank	0.000	04:49
Subject Sample # 1	0.000	04:50
Breath Volume	2.539	04:50
Air Blank	0.000	04:50
Air Blank	0.000	04:52
Subject Sample # 2	0.000	04:53
Breath Volume	1.523	04:53
Air Blank	0.000	04:53
Control Test	0.078	04:53
Air Blank	0.000	04:54
Diagnostics Check	OK	04:54

Data Download Date/Time: 08/10/2016 20:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMMONS,ANGELA,V
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HERD,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 11:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 12:50

SUBJECT'S LAST NAME: FAROTTO

DATE/TIME OF TEST: 07/15/2016 13:16:54

Diagnostics Check	OK	13:11
Air Blank	0.000	13:11
Control Test	0.079	13:11
Air Blank	0.000	13:12
Subject Sample # 1	0.257	13:12
Breath Volume	2.101	13:12
Air Blank	0.000	13:13
Air Blank	0.000	13:15
Subject Sample # 2	0.263	13:15
Breath Volume	2.089	13:15
Air Blank	0.000	13:16
Control Test	0.077	13:16
Air Blank	0.000	13:16
Diagnostics Check	OK	13:16

Data Download Date/Time: 08/10/2016 20:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY,PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:18

SUBJECT'S LAST NAME: CASTANEDA

DATE/TIME OF TEST: 07/16/2016 02:46:46

Diagnostics Check	OK	02:38
Air Blank	0.000	02:39
Control Test	0.079	02:39
Air Blank	0.000	02:39
Subject Sample # 1	0.073	02:42
Breath Volume	1.535	02:42
Air Blank	0.000	02:42
Air Blank	0.000	02:44
Subject Sample # 2	0.072	02:45
Breath Volume	1.558	02:45
Air Blank	0.000	02:45
Control Test	0.078	02:46
Air Blank	0.000	02:46
Diagnostics Check	OK	02:46

Data Download Date/Time: 08/10/2016 20:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,REBECCA,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GILBERT,T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 15:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:40

SUBJECT'S LAST NAME: HIGGINS

DATE/TIME OF TEST: 07/16/2016 17:03:21

Diagnostics Check	OK	17:00	
Air Blank	0.000	17:01	
Control Test	0.078	17:01	
Air Blank	0.000	17:01	
Subject Sample	# 1	17:02	Subject Test Refused
Breath Volume	0.000	17:02	
Air Blank	0.000	17:02	
Control Test	0.078	17:02	
Air Blank	0.000	17:03	
Diagnostics Check	OK	17:03	

Data Download Date/Time: 08/10/2016 20:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MOSS,J
ARRESTING OFFICER AGENCY: Temple Terrace PD
LAST AGENCY INSPECTION: 07/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 18:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:37

SUBJECT'S LAST NAME: POWERS II

DATE/TIME OF TEST: 07/16/2016 19:04:12

Diagnostics Check	OK	18:57
Air Blank	0.000	18:58
Control Test	0.079	18:58
Air Blank	0.000	18:59
Subject Sample # 1	0.086	18:59
Breath Volume	1.480	18:59
Air Blank	0.000	19:00
Air Blank	0.000	19:02
Subject Sample # 2	0.094	19:02
Breath Volume	1.988	19:02
Air Blank	0.000	19:03
Control Test	0.078	19:03
Air Blank	0.000	19:04
Diagnostics Check	OK	19:04

Data Download Date/Time: 08/10/2016 20:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARLAUG,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:06

SUBJECT'S LAST NAME: COSPER

DATE/TIME OF TEST: 07/16/2016 23:33:22

Diagnostics Check	OK	23:26
Air Blank	0.000	23:27
Control Test	0.079	23:27
Air Blank	0.000	23:28
Subject Sample # 1	0.145	23:28
Breath Volume	1.386	23:28
Air Blank	0.000	23:28
Air Blank	0.000	23:30
Subject Sample # 2	0.150	23:31
Breath Volume	1.824	23:31
Air Blank	0.000	23:32
Control Test	0.078	23:32
Air Blank	0.000	23:33
Diagnostics Check	OK	23:33

Data Download Date/Time: 08/10/2016 20:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PALFY,R
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:18

SUBJECT'S LAST NAME: GLASGOW

DATE/TIME OF TEST: 07/16/2016 23:47:27

Diagnostics Check	OK	23:40
Air Blank	0.000	23:40
Control Test	0.078	23:40
Air Blank	0.000	23:41
Subject Sample # 1	0.110	23:42
Breath Volume	1.242	23:42
Air Blank	0.000	23:42
Air Blank	0.000	23:44
Subject Sample # 2	0.115	23:45
Breath Volume	2.597	23:45
Air Blank	0.000	23:46
Control Test	0.078	23:46
Air Blank	0.000	23:47
Diagnostics Check	OK	23:47

Data Download Date/Time: 08/10/2016 20:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GONZALEZ,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:42

SUBJECT'S LAST NAME: FLORES RODRIGUEZ

DATE/TIME OF TEST: 07/17/2016 04:15:44

Diagnostics Check		OK	04:02	
Air Blank		0.000	04:03	
Control Test		0.079	04:03	
Air Blank		0.000	04:04	
Subject Sample	# 1	0.182	04:07	Volume Not Met
Breath Volume		1.074	04:07	
Air Blank		0.000	04:07	
Air Blank		0.000	04:09	
Subject Sample	# 2	0.208	04:10	
Breath Volume		2.996	04:10	
Air Blank		0.000	04:11	
Air Blank		0.000	04:13	
Subject Sample	# 3	0.189	04:14	
Breath Volume		1.164	04:14	
Air Blank		0.000	04:14	
Control Test		0.078	04:15	
Air Blank		0.000	04:15	
Diagnostics Check		OK	04:15	

Data Download Date/Time: 08/10/2016 20:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRABAND,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:38

SUBJECT'S LAST NAME: HOOVER

DATE/TIME OF TEST: 07/17/2016 05:06:27

Diagnostics Check	OK	04:58
Air Blank	0.000	04:59
Control Test	0.078	04:59
Air Blank	0.000	05:00
Subject Sample # 1	0.125	05:01
Breath Volume	1.269	05:01
Air Blank	0.000	05:02
Air Blank	0.000	05:04
Subject Sample # 2	0.127	05:04
Breath Volume	2.207	05:04
Air Blank	0.000	05:05
Control Test	0.078	05:05
Air Blank	0.000	05:06
Diagnostics Check	OK	05:06

Data Download Date/Time: 08/10/2016 20:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SIMPKINS,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 05:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:19

SUBJECT'S LAST NAME: PARR

DATE/TIME OF TEST: 07/17/2016 06:45:33

Diagnostics Check	OK	06:39
Air Blank	0.000	06:39
Control Test	0.078	06:39
Air Blank	0.000	06:40
Subject Sample # 1	0.147	06:40
Breath Volume	1.292	06:40
Air Blank	0.000	06:41
Air Blank	0.000	06:43
Subject Sample # 2	0.160	06:44
Breath Volume	2.363	06:44
Air Blank	0.000	06:44
Control Test	0.078	06:44
Air Blank	0.000	06:45
Diagnostics Check	OK	06:45

Data Download Date/Time: 08/10/2016 20:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:25

SUBJECT'S LAST NAME: LOVETT

DATE/TIME OF TEST: 07/17/2016 20:53:30

Diagnostics Check	OK	20:45
Air Blank	0.000	20:46
Control Test	0.079	20:46
Air Blank	0.000	20:47
Subject Sample # 1	0.258	20:48
Breath Volume	1.281	20:48
Air Blank	0.000	20:49
Air Blank	0.000	20:51
Subject Sample # 2	0.266	20:52
Breath Volume	1.523	20:52
Air Blank	0.000	20:52
Control Test	0.077	20:52
Air Blank	0.000	20:53
Diagnostics Check	OK	20:53

Data Download Date/Time: 08/10/2016 20:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PORTMAN,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:11

SUBJECT'S LAST NAME: CABALLERO SILVERIO

DATE/TIME OF TEST: 07/18/2016 03:37:48

Diagnostics Check	OK	03:31
Air Blank	0.000	03:32
Control Test	0.079	03:32
Air Blank	0.000	03:33
Subject Sample # 1	0.064	03:33
Breath Volume	1.953	03:33
Air Blank	0.000	03:34
Air Blank	0.000	03:35
Subject Sample # 2	0.072	03:36
Breath Volume	2.316	03:36
Air Blank	0.000	03:36
Control Test	0.078	03:37
Air Blank	0.000	03:37
Diagnostics Check	OK	03:37

Data Download Date/Time: 08/10/2016 20:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: STEDMAN

DATE/TIME OF TEST: 07/19/2016 01:55:12

Diagnostics Check	OK	01:45
Air Blank	0.000	01:45
Control Test	0.079	01:46
Air Blank	0.000	01:46
Subject Sample # 1	0.200	01:47
Breath Volume	1.351	01:47
Air Blank	0.000	01:47
Air Blank	0.000	01:49
Subject Sample # 2	0.238	01:50
Breath Volume	3.273	01:50
Air Blank	0.000	01:51
Air Blank	0.000	01:52
Subject Sample # 3	0.253	01:53
Breath Volume	4.933	01:53
Air Blank	0.000	01:54
Control Test	0.079	01:54
Air Blank	0.000	01:55
Diagnostics Check	OK	01:55

Data Download Date/Time: 08/10/2016 20:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY, PTRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DAWSY, B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:26

SUBJECT'S LAST NAME: BARNWELL

DATE/TIME OF TEST: 07/19/2016 22:53:08

Diagnostics Check	OK	22:46
Air Blank	0.000	22:47
Control Test	0.079	22:47
Air Blank	0.000	22:48
Subject Sample # 1	0.214	22:48
Breath Volume	1.675	22:48
Air Blank	0.000	22:49
Air Blank	0.000	22:51
Subject Sample # 2	0.218	22:51
Breath Volume	1.761	22:51
Air Blank	0.000	22:52
Control Test	0.079	22:52
Air Blank	0.000	22:53
Diagnostics Check	OK	22:53

Data Download Date/Time: 08/10/2016 20:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GONZALEZ, J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: ESTIVERNE

DATE/TIME OF TEST: 07/20/2016 04:37:45

Diagnostics Check	OK	04:30
Air Blank	0.000	04:30
Control Test	0.079	04:31
Air Blank	0.000	04:31
Subject Sample # 1	0.103	04:33
Breath Volume	1.378	04:33
Air Blank	0.000	04:33
Air Blank	0.000	04:35
Subject Sample # 2	0.100	04:36
Breath Volume	1.304	04:36
Air Blank	0.000	04:36
Control Test	0.079	04:37
Air Blank	0.000	04:37
Diagnostics Check	OK	04:37

Data Download Date/Time: 08/10/2016 20:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY, PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS, S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:51

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 07/21/2016 02:18:10

Diagnostics Check	OK	02:11
Air Blank	0.000	02:12
Control Test	0.078	02:12
Air Blank	0.000	02:12
Subject Sample # 1	0.161	02:13
Breath Volume	3.250	02:13
Air Blank	0.000	02:14
Air Blank	0.000	02:15
Subject Sample # 2	0.160	02:16
Breath Volume	3.621	02:16
Air Blank	0.000	02:17
Control Test	0.078	02:17
Air Blank	0.000	02:18
Diagnostics Check	OK	02:18

Data Download Date/Time: 08/10/2016 20:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK,CASEY,S
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DEJESUS-COLON,S
ARRESTING OFFICER AGENCY: Temple Terrace PD
LAST AGENCY INSPECTION: 07/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: ARIAS

DATE/TIME OF TEST: 07/22/2016 04:31:50

Diagnostics Check	OK	04:25
Air Blank	0.000	04:26
Control Test	0.079	04:26
Air Blank	0.000	04:27
Subject Sample # 1	0.143	04:27
Breath Volume	1.984	04:27
Air Blank	0.000	04:28
Air Blank	0.000	04:29
Subject Sample # 2	0.144	04:30
Breath Volume	1.625	04:30
Air Blank	0.000	04:30
Control Test	0.079	04:31
Air Blank	0.000	04:31
Diagnostics Check	OK	04:31

Data Download Date/Time: 08/10/2016 20:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNYDER,MELONIE,E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SHANNON,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 06:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:14

SUBJECT'S LAST NAME: LAMBERT

DATE/TIME OF TEST: 07/22/2016 07:40:36

Diagnostics Check	OK	07:34
Air Blank	0.000	07:34
Control Test	0.078	07:34
Air Blank	0.000	07:35
Subject Sample # 1	0.233	07:36
Breath Volume	3.906	07:36
Air Blank	0.000	07:36
Air Blank	0.000	07:38
Subject Sample # 2	0.238	07:39
Breath Volume	3.496	07:39
Air Blank	0.000	07:39
Control Test	0.078	07:40
Air Blank	0.000	07:40
Diagnostics Check	OK	07:40

Data Download Date/Time: 08/10/2016 20:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK,CASEY,S
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DEJESUS,V
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:42

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 07/23/2016 01:11:10

Diagnostics Check		OK	01:04
Air Blank		0.000	01:04
Control Test		0.079	01:04
Air Blank		0.000	01:05
Subject Sample	# 1	0.118	01:05
Breath Volume		1.386	01:05
Air Blank		0.000	01:06
Air Blank		0.000	01:08
Subject Sample	# 2	0.122	01:09
Breath Volume		1.406	01:09
Air Blank		0.000	01:10
Control Test		0.078	01:10
Air Blank		0.000	01:11
Diagnostics Check		OK	01:11

Data Download Date/Time: 08/10/2016 20:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BELMONTE,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: WEISS

DATE/TIME OF TEST: 07/23/2016 02:02:55

Diagnostics Check	OK	01:55
Air Blank	0.000	01:56
Control Test	0.078	01:56
Air Blank	0.000	01:57
Subject Sample # 1	0.052	01:58
Breath Volume	3.500	01:58
Air Blank	0.000	01:59
Air Blank	0.000	02:01
Subject Sample # 2	0.052	02:01
Breath Volume	2.425	02:01
Air Blank	0.000	02:02
Control Test	0.079	02:02
Air Blank	0.000	02:02
Diagnostics Check	OK	02:02

Data Download Date/Time: 08/10/2016 20:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PECYLAK,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:43

SUBJECT'S LAST NAME: DAVIDSON

DATE/TIME OF TEST: 07/23/2016 03:10:59

Diagnostics Check	OK	03:04
Air Blank	0.000	03:04
Control Test	0.080	03:04
Air Blank	0.000	03:05
Subject Sample # 1	0.200	03:06
Breath Volume	1.730	03:06
Air Blank	0.000	03:06
Air Blank	0.000	03:08
Subject Sample # 2	0.206	03:09
Breath Volume	2.242	03:09
Air Blank	0.000	03:10
Control Test	0.078	03:10
Air Blank	0.000	03:10
Diagnostics Check	OK	03:10

Data Download Date/Time: 08/10/2016 20:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK,CASEY,S
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SIERRA,P
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:22

SUBJECT'S LAST NAME: MCCURDY

DATE/TIME OF TEST: 07/23/2016 03:46:39

Diagnostics Check	OK	03:43	
Air Blank	0.000	03:44	
Control Test	0.080	03:44	
Air Blank	0.000	03:44	
Subject Sample	# 1	03:45	RFI Detect
Breath Volume	0.000	03:45	
Air Blank	0.000	03:45	
Control Test	0.078	03:46	
Air Blank	0.000	03:46	
Diagnostics Check	OK	03:46	

Data Download Date/Time: 08/10/2016 20:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK,CASEY,S
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SIERRA,P
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:22

SUBJECT'S LAST NAME: MCCURDY

DATE/TIME OF TEST: 07/23/2016 03:55:08

Diagnostics Check		OK	03:48
Air Blank		0.000	03:49
Control Test		0.078	03:49
Air Blank		0.000	03:50
Subject Sample	# 1	0.132	03:50
Breath Volume		1.484	03:50
Air Blank		0.000	03:51
Air Blank		0.000	03:53
Subject Sample	# 2	0.138	03:53
Breath Volume		1.941	03:53
Air Blank		0.000	03:54
Control Test		0.078	03:54
Air Blank		0.000	03:55
Diagnostics Check		OK	03:55

Data Download Date/Time: 08/10/2016 20:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARLAUG,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: WRIGHT JR

DATE/TIME OF TEST: 07/23/2016 04:37:02

Diagnostics Check	OK	04:30
Air Blank	0.000	04:31
Control Test	0.079	04:31
Air Blank	0.000	04:31
Subject Sample # 1	0.129	04:32
Breath Volume	1.718	04:32
Air Blank	0.000	04:33
Air Blank	0.000	04:35
Subject Sample # 2	0.128	04:35
Breath Volume	2.542	04:35
Air Blank	0.000	04:36
Control Test	0.078	04:36
Air Blank	0.000	04:36
Diagnostics Check	OK	04:37

Data Download Date/Time: 08/10/2016 20:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WOODEN,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:41

SUBJECT'S LAST NAME: CARTER

DATE/TIME OF TEST: 07/23/2016 05:09:19

Diagnostics Check		OK	05:02
Air Blank		0.000	05:02
Control Test		0.079	05:03
Air Blank		0.000	05:03
Subject Sample	# 1	0.205	05:04
Breath Volume		2.750	05:04
Air Blank		0.000	05:05
Air Blank		0.000	05:07
Subject Sample	# 2	0.207	05:07
Breath Volume		3.050	05:07
Air Blank		0.000	05:08
Control Test		0.078	05:08
Air Blank		0.000	05:09
Diagnostics Check		OK	05:09

Data Download Date/Time: 08/10/2016 20:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DE JESUS, V
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 19:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:48

SUBJECT'S LAST NAME: FUENTES LOPEZ

DATE/TIME OF TEST: 07/23/2016 20:15:17

Diagnostics Check	OK	20:08
Air Blank	0.000	20:09
Control Test	0.079	20:09
Air Blank	0.000	20:10
Subject Sample # 1	0.210	20:10
Breath Volume	2.308	20:10
Air Blank	0.000	20:11
Air Blank	0.000	20:13
Subject Sample # 2	0.213	20:13
Breath Volume	1.253	20:13
Air Blank	0.000	20:14
Control Test	0.079	20:14
Air Blank	0.000	20:15
Diagnostics Check	OK	20:15

Data Download Date/Time: 08/10/2016 20:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY,PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRABAND,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 19:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: CARMENATE MEDINA

DATE/TIME OF TEST: 07/23/2016 21:46:59

Diagnostics Check	OK	21:40
Air Blank	0.000	21:41
Control Test	0.080	21:41
Air Blank	0.000	21:41
Subject Sample # 1	0.042	21:42
Breath Volume	1.199	21:42
Air Blank	0.000	21:43
Air Blank	0.000	21:44
Subject Sample # 2	0.039	21:45
Breath Volume	1.585	21:45
Air Blank	0.000	21:46
Control Test	0.079	21:46
Air Blank	0.000	21:46
Diagnostics Check	OK	21:46

Data Download Date/Time: 08/10/2016 20:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GONZALEZ,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:40

SUBJECT'S LAST NAME: CLARK

DATE/TIME OF TEST: 07/24/2016 05:08:32

Diagnostics Check		OK	05:02
Air Blank		0.000	05:02
Control Test		0.079	05:03
Air Blank		0.000	05:03
Subject Sample	# 1	0.158	05:04
Breath Volume		2.550	05:04
Air Blank		0.000	05:04
Air Blank		0.000	05:06
Subject Sample	# 2	0.154	05:07
Breath Volume		1.644	05:07
Air Blank		0.000	05:07
Control Test		0.078	05:07
Air Blank		0.000	05:08
Diagnostics Check		OK	05:08

Data Download Date/Time: 08/10/2016 20:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNIDER, P
ARRESTING OFFICER AGENCY: Plant City PD
LAST AGENCY INSPECTION: 07/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 05:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:12

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 07/24/2016 06:39:40

Diagnostics Check		OK	06:32
Air Blank		0.000	06:32
Control Test		0.079	06:33
Air Blank		0.000	06:33
Subject Sample	# 1	0.130	06:35
Breath Volume		1.847	06:35
Air Blank		0.000	06:35
Air Blank		0.000	06:37
Subject Sample	# 2	0.125	06:38
Breath Volume		1.828	06:38
Air Blank		0.000	06:38
Control Test		0.078	06:39
Air Blank		0.000	06:39
Diagnostics Check		OK	06:39

Data Download Date/Time: 08/10/2016 20:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY, PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DAWSY, B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:33

SUBJECT'S LAST NAME: RAMIREZ MAYA

DATE/TIME OF TEST: 07/25/2016 02:01:25

Diagnostics Check	OK	01:54
Air Blank	0.000	01:55
Control Test	0.079	01:55
Air Blank	0.000	01:56
Subject Sample # 1	0.214	01:56
Breath Volume	2.582	01:56
Air Blank	0.000	01:57
Air Blank	0.000	01:59
Subject Sample # 2	0.210	01:59
Breath Volume	3.378	01:59
Air Blank	0.000	02:00
Control Test	0.078	02:00
Air Blank	0.000	02:01
Diagnostics Check	OK	02:01

Data Download Date/Time: 08/10/2016 20:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PECYLAK, M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:21

SUBJECT'S LAST NAME: RICHTER

DATE/TIME OF TEST: 07/29/2016 03:48:08

Diagnostics Check		OK	03:41
Air Blank		0.000	03:42
Control Test		0.079	03:42
Air Blank		0.000	03:43
Subject Sample	# 1	0.291	03:43
Breath Volume		1.796	03:43
Air Blank		0.000	03:44
Air Blank		0.000	03:46
Subject Sample	# 2	0.279	03:46
Breath Volume		1.687	03:46
Air Blank		0.000	03:47
Control Test		0.079	03:47
Air Blank		0.000	03:48
Diagnostics Check		OK	03:48

Data Download Date/Time: 08/10/2016 20:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMMONS,ANGELA,V
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DAVIS,LARRY
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 14:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:31

SUBJECT'S LAST NAME: ZOCCOLI

DATE/TIME OF TEST: 07/29/2016 16:03:20

Diagnostics Check	OK	15:51
Air Blank	0.000	15:52
Control Test	0.079	15:52
Air Blank	0.000	15:52
Subject Sample # 1	0.319	15:54
Breath Volume	1.707	15:54
Air Blank	0.000	15:55
Air Blank	0.000	15:56
Subject Sample # 2	0.298	15:57
Breath Volume	1.417	15:57
Air Blank	0.000	15:58
Air Blank	0.000	16:00
Subject Sample # 3	0.306	16:01
Breath Volume	1.496	16:01
Air Blank	0.000	16:02
Control Test	0.078	16:02
Air Blank	0.000	16:03
Diagnostics Check	OK	16:03

Data Download Date/Time: 08/10/2016 20:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY,PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:34

SUBJECT'S LAST NAME: HULSEY

DATE/TIME OF TEST: 07/30/2016 02:05:07

Diagnostics Check	OK	01:54
Air Blank	0.000	01:55
Control Test	0.079	01:55
Air Blank	0.000	01:55
Subject Sample # 1	0.267	01:56
Breath Volume	1.582	01:56
Air Blank	0.000	01:57
Air Blank	0.000	01:59
Subject Sample # 2	0.237	01:59
Breath Volume	1.421	01:59
Air Blank	0.000	02:00
Air Blank	0.000	02:02
Subject Sample # 3	0.236	02:03
Breath Volume	1.312	02:03
Air Blank	0.000	02:04
Control Test	0.078	02:04
Air Blank	0.000	02:05
Diagnostics Check	OK	02:05

Data Download Date/Time: 08/10/2016 20:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY,PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VAN-TREESE,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:21

SUBJECT'S LAST NAME: MAZUR

DATE/TIME OF TEST: 07/30/2016 04:48:22

Diagnostics Check	OK	04:42
Air Blank	0.000	04:42
Control Test	0.079	04:42
Air Blank	0.000	04:43
Subject Sample # 1	0.000	04:43
Breath Volume	2.890	04:43
Air Blank	0.000	04:44
Air Blank	0.000	04:46
Subject Sample # 2	0.000	04:47
Breath Volume	2.089	04:47
Air Blank	0.000	04:47
Control Test	0.079	04:47
Air Blank	0.000	04:48
Diagnostics Check	OK	04:48

Data Download Date/Time: 08/10/2016 20:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LYON,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: BOGUE

DATE/TIME OF TEST: 07/31/2016 03:12:52

Diagnostics Check		OK	03:05
Air Blank		0.000	03:06
Control Test		0.079	03:06
Air Blank		0.000	03:06
Subject Sample	# 1	0.177	03:08
Breath Volume		2.699	03:08
Air Blank		0.000	03:08
Air Blank		0.000	03:10
Subject Sample	# 2	0.164	03:11
Breath Volume		1.902	03:11
Air Blank		0.000	03:11
Control Test		0.077	03:12
Air Blank		0.000	03:12
Diagnostics Check		OK	03:12

Data Download Date/Time: 08/10/2016 20:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:40

SUBJECT'S LAST NAME: OBREGON JR

DATE/TIME OF TEST: 07/31/2016 05:06:32

Diagnostics Check		OK	05:00
Air Blank		0.000	05:00
Control Test		0.079	05:00
Air Blank		0.000	05:01
Subject Sample	# 1	0.193	05:02
Breath Volume		2.796	05:02
Air Blank		0.000	05:02
Air Blank		0.000	05:04
Subject Sample	# 2	0.189	05:05
Breath Volume		2.992	05:05
Air Blank		0.000	05:05
Control Test		0.079	05:05
Air Blank		0.000	05:06
Diagnostics Check		OK	05:06

Data Download Date/Time: 08/10/2016 20:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LARA,A
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:19

SUBJECT'S LAST NAME: TWEEDALE

DATE/TIME OF TEST: 07/31/2016 23:45:40

Diagnostics Check	OK	23:39
Air Blank	0.000	23:39
Control Test	0.080	23:39
Air Blank	0.000	23:40
Subject Sample # 1	0.096	23:41
Breath Volume	2.726	23:41
Air Blank	0.000	23:41
Air Blank	0.000	23:43
Subject Sample # 2	0.099	23:44
Breath Volume	1.914	23:44
Air Blank	0.000	23:44
Control Test	0.079	23:45
Air Blank	0.000	23:45
Diagnostics Check	OK	23:45

Data Download Date/Time: 08/10/2016 20:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSS,ROBERT,R
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: MOSS,ROBERT
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 06/17/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1752856
DRY GAS EXPIRY DATE: 10/13/2017
VIOLATION CODE: DUI
ARREST TIME: 17:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:55

SUBJECT'S LAST NAME: ASHCRAFT III

DATE/TIME OF TEST: 07/01/2016 18:21:34

Diagnostics Check	OK	18:18	
Air Blank	0.000	18:18	
Control Test	0.080	18:19	
Air Blank	0.000	18:19	
Subject Sample	# 1	18:20	Interferent Detect
Breath Volume	3.019	18:20	
Air Blank	0.000	18:20	
Control Test	0.080	18:20	
Air Blank	0.000	18:21	
Diagnostics Check	OK	18:21	

Data Download Date/Time: 07/26/2016 10:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSS,ROBERT,R
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: MOSS,ROBERT
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 06/17/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1752856
DRY GAS EXPIRY DATE: 10/13/2017
VIOLATION CODE: DUI
ARREST TIME: 17:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:55

SUBJECT'S LAST NAME: ASHCRAFT III

DATE/TIME OF TEST: 07/01/2016 18:26:34

Diagnostics Check	OK	18:23	
Air Blank	0.000	18:23	
Control Test	0.080	18:24	
Air Blank	0.000	18:24	
Subject Sample	# 1	18:25	Interferent Detect
Breath Volume	1.699	18:25	
Air Blank	0.000	18:25	
Control Test	0.080	18:25	
Air Blank	0.000	18:26	
Diagnostics Check	OK	18:26	

Data Download Date/Time: 07/26/2016 10:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSS,ROBERT,R
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: MOSS,ROBERT
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 06/17/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1752856
DRY GAS EXPIRY DATE: 10/13/2017
VIOLATION CODE: DUI
ARREST TIME: 17:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:55

SUBJECT'S LAST NAME: ASHCRAFT III

DATE/TIME OF TEST: 07/01/2016 18:37:19

Diagnostics Check	OK	18:34	
Air Blank	0.000	18:34	
Control Test	0.075	18:34	
Air Blank	0.000	18:35	
Subject Sample	# 1	18:35	Interferent Detect
Breath Volume	3.542	18:35	
Air Blank	0.000	18:36	
Control Test	0.080	18:36	
Air Blank	0.000	18:37	
Diagnostics Check	OK	18:37	

Data Download Date/Time: 07/26/2016 10:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,STEPHANIE,N
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: BROOKS,BLAKE
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 06/17/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1752856
DRY GAS EXPIRY DATE: 10/13/2017
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: ROSS

DATE/TIME OF TEST: 07/02/2016 03:57:04

Diagnostics Check	OK	03:50
Air Blank	0.000	03:51
Control Test	0.080	03:51
Air Blank	0.000	03:52
Subject Sample # 1	0.122	03:52
Breath Volume	3.347	03:52
Air Blank	0.000	03:53
Air Blank	0.000	03:55
Subject Sample # 2	0.124	03:55
Breath Volume	2.906	03:55
Air Blank	0.000	03:56
Control Test	0.081	03:56
Air Blank	0.000	03:56
Diagnostics Check	OK	03:57

Data Download Date/Time: 07/26/2016 10:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSS,ROBERT,R
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: MOSS,ROBERT
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 06/17/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1752856
DRY GAS EXPIRY DATE: 10/13/2017
VIOLATION CODE: DUI
ARREST TIME: 00:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:07

SUBJECT'S LAST NAME: ROBERTS

DATE/TIME OF TEST: 07/03/2016 00:45:05

Diagnostics Check	OK	00:37
Air Blank	0.000	00:37
Control Test	0.081	00:38
Air Blank	0.000	00:38
Subject Sample # 1	0.179	00:40
Breath Volume	2.941	00:40
Air Blank	0.000	00:41
Air Blank	0.000	00:43
Subject Sample # 2	0.172	00:43
Breath Volume	2.734	00:43
Air Blank	0.000	00:44
Control Test	0.080	00:44
Air Blank	0.000	00:45
Diagnostics Check	OK	00:45

Data Download Date/Time: 07/26/2016 10:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROOKS, BLAKE, T
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: BROOKS, BLAKE
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 06/17/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1752856
DRY GAS EXPIRY DATE: 10/13/2017
VIOLATION CODE: DUI
ARREST TIME: 00:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:19

SUBJECT'S LAST NAME: STROHRIGL

DATE/TIME OF TEST: 07/05/2016 00:47:49

Diagnostics Check	OK	00:41
Air Blank	0.000	00:41
Control Test	0.080	00:42
Air Blank	0.000	00:42
Subject Sample # 1	0.200	00:43
Breath Volume	1.414	00:43
Air Blank	0.000	00:44
Air Blank	0.000	00:45
Subject Sample # 2	0.210	00:46
Breath Volume	1.699	00:46
Air Blank	0.000	00:46
Control Test	0.079	00:47
Air Blank	0.000	00:47
Diagnostics Check	OK	00:47

Data Download Date/Time: 07/26/2016 10:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROOKS, BLAKE, T
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: ARAFA, YUSUF
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 06/17/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1752856
DRY GAS EXPIRY DATE: 10/13/2017
VIOLATION CODE: DUI
ARREST TIME: 02:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: BIRCHFIELD

DATE/TIME OF TEST: 07/05/2016 03:02:08

Diagnostics Check	OK	02:55
Air Blank	0.000	02:56
Control Test	0.080	02:56
Air Blank	0.000	02:57
Subject Sample # 1	0.223	02:57
Breath Volume	1.851	02:57
Air Blank	0.000	02:58
Air Blank	0.000	03:00
Subject Sample # 2	0.208	03:00
Breath Volume	1.281	03:00
Air Blank	0.000	03:01
Control Test	0.079	03:01
Air Blank	0.000	03:02
Diagnostics Check	OK	03:02

Data Download Date/Time: 07/26/2016 10:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROOKS, BLAKE, T
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: BROOKS, BLAKE
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 06/17/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1752856
DRY GAS EXPIRY DATE: 10/13/2017
VIOLATION CODE: DUI
ARREST TIME: 04:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:21

SUBJECT'S LAST NAME: BOBER

DATE/TIME OF TEST: 07/07/2016 04:54:50

Diagnostics Check	OK	04:44
Air Blank	0.000	04:44
Control Test	0.080	04:45
Air Blank	0.000	04:45
Subject Sample # 1	0.206	04:48
Breath Volume	1.433	04:48
Air Blank	0.000	04:49
Air Blank	0.000	04:51
Subject Sample # 2	0.194	04:53
Breath Volume	1.597	04:53
Air Blank	0.000	04:53
Control Test	0.078	04:54
Air Blank	0.000	04:54
Diagnostics Check	OK	04:54

Data Download Date/Time: 07/26/2016 10:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSS,ROBERT,R
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: MOSS,ROBERT
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 06/17/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1752856
DRY GAS EXPIRY DATE: 10/13/2017
VIOLATION CODE: DUI
ARREST TIME: 21:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:55

SUBJECT'S LAST NAME: HANNAH

DATE/TIME OF TEST: 07/19/2016 22:21:59

Diagnostics Check	OK	22:15
Air Blank	0.000	22:16
Control Test	0.080	22:16
Air Blank	0.000	22:17
Subject Sample # 1	0.313	22:17
Breath Volume	2.988	22:17
Air Blank	0.000	22:18
Air Blank	0.000	22:20
Subject Sample # 2	0.310	22:20
Breath Volume	3.421	22:20
Air Blank	0.000	22:21
Control Test	0.080	22:21
Air Blank	0.000	22:21
Diagnostics Check	OK	22:21

Data Download Date/Time: 07/26/2016 10:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROOKS, BLAKE, T
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: BROOKS, BLAKE
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 06/17/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1752856
DRY GAS EXPIRY DATE: 10/13/2017
VIOLATION CODE: DUI
ARREST TIME: 23:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: OWEN

DATE/TIME OF TEST: 07/19/2016 23:53:58

Diagnostics Check	OK	23:51	
Air Blank	0.000	23:52	
Control Test	0.081	23:52	
Air Blank	0.000	23:52	
Subject Sample	# 1	23:53	RFI Detect
Breath Volume	0.000	23:53	
Air Blank		23:53	RFI Detect
Air Blank	0.000	23:53	

Data Download Date/Time: 07/26/2016 10:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROOKS, BLAKE, T
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: BROOKS, BLAKE
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 06/17/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1752856
DRY GAS EXPIRY DATE: 10/13/2017
VIOLATION CODE: DUI
ARREST TIME: 23:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: OWEN

DATE/TIME OF TEST: 07/20/2016 00:02:19

Diagnostics Check	OK	23:55
Air Blank	0.000	23:56
Control Test	0.081	23:56
Air Blank	0.000	23:56
Subject Sample # 2	0.068	23:57
Breath Volume	1.152	23:57
Air Blank	0.000	23:58
Air Blank	0.000	23:59
Subject Sample # 1	0.078	00:00
Breath Volume	1.769	00:00
Air Blank	0.000	00:01
Control Test	0.080	00:01
Air Blank	0.000	00:02
Diagnostics Check	OK	00:02

Data Download Date/Time: 07/26/2016 10:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSS,ROBERT,R
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: MOSS,ROBERT
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 06/17/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1752856
DRY GAS EXPIRY DATE: 10/13/2017
VIOLATION CODE: DUI
ARREST TIME: 21:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: COKER

DATE/TIME OF TEST: 07/25/2016 23:13:12

Diagnostics Check	OK	23:07
Air Blank	0.000	23:07
Control Test	0.081	23:07
Air Blank	0.000	23:08
Subject Sample # 1	0.175	23:08
Breath Volume	1.492	23:08
Air Blank	0.000	23:09
Air Blank	0.000	23:11
Subject Sample # 2	0.192	23:11
Breath Volume	1.949	23:11
Air Blank	0.000	23:12
Control Test	0.080	23:12
Air Blank	0.000	23:13
Diagnostics Check	OK	23:13

Data Download Date/Time: 07/26/2016 10:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: LYNN HAVEN PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KELLY,RALPH,E
BREATH TEST OPERATOR AGENCY: Lynn Haven PD
ARRESTING OFFICER: SHIPES,RYAN
ARRESTING OFFICER AGENCY: Lynn Haven PD
LAST AGENCY INSPECTION: 06/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29614080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 21:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: GREGORY

DATE/TIME OF TEST: 07/23/2016 22:49:43

Diagnostics Check	OK	22:41	
Air Blank	0.000	22:42	
Control Test	0.075	22:42	
Air Blank	0.000	22:42	
Subject Sample # 1	0.218	22:45	
Breath Volume	1.992	22:45	
Air Blank	0.000	22:45	
Air Blank	0.000	22:47	
Subject Sample # 2	0.230	22:48	
Breath Volume	3.679	22:48	
Air Blank	0.000	22:48	
Control Test	0.074	22:49	Control Outside Tolerance
Air Blank	0.000	22:49	

Data Download Date/Time: 07/27/2016 18:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: LYNN HAVEN PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NUSBAUM,K,
BREATH TEST OPERATOR AGENCY: Lynn Haven PD
ARRESTING OFFICER: NUSBAUM,K
ARRESTING OFFICER AGENCY: Lynn Haven PD
LAST AGENCY INSPECTION: 07/27/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29614080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 01:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:13

SUBJECT'S LAST NAME: AGAR

DATE/TIME OF TEST: 07/29/2016 01:54:40

Diagnostics Check	OK	01:51	
Air Blank	0.000	01:52	
Control Test	0.076	01:52	
Air Blank	0.000	01:53	
Subject Sample	# 1	01:53	Subject Test Refused
Breath Volume	0.000	01:53	
Air Blank	0.000	01:53	
Control Test	0.077	01:54	
Air Blank	0.000	01:54	
Diagnostics Check	OK	01:54	

Data Download Date/Time: 08/29/2016 23:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,DANNY,W
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: GRAY,DANNY
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 06/08/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:14

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 07/02/2016 03:42:13

Diagnostics Check	OK	03:36
Air Blank	0.000	03:36
Control Test	0.080	03:37
Air Blank	0.000	03:37
Subject Sample # 1	0.137	03:37
Breath Volume	3.226	03:37
Air Blank	0.000	03:38
Air Blank	0.000	03:40
Subject Sample # 2	0.136	03:40
Breath Volume	3.164	03:40
Air Blank	0.000	03:41
Control Test	0.078	03:41
Air Blank	0.000	03:42
Diagnostics Check	OK	03:42

Data Download Date/Time: 07/06/2016 10:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZARRILLO,N,
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: WILSON,P
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 06/08/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: NAVAS

DATE/TIME OF TEST: 07/02/2016 22:56:57

Diagnostics Check	OK	22:55	
Air Blank	0.000	22:56	
Control Test	0.080	22:56	
Air Blank		22:56	RFI Detect
Air Blank	0.000	22:56	

Data Download Date/Time: 07/06/2016 10:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZARRILLO,N,
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: WILSON,P
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 06/08/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: NAVAS

DATE/TIME OF TEST: 07/02/2016 23:04:52

Diagnostics Check	OK	22:58
Air Blank	0.000	22:59
Control Test	0.081	22:59
Air Blank	0.000	22:59
Subject Sample # 1	0.127	23:00
Breath Volume	2.523	23:00
Air Blank	0.000	23:00
Air Blank	0.000	23:02
Subject Sample # 2	0.129	23:03
Breath Volume	2.296	23:03
Air Blank	0.000	23:03
Control Test	0.080	23:04
Air Blank	0.000	23:04
Diagnostics Check	OK	23:04

Data Download Date/Time: 07/06/2016 10:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARSON,JAMES,A
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: CARSON,JAMES
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 06/08/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:24

SUBJECT'S LAST NAME: DRISCOLL

DATE/TIME OF TEST: 07/03/2016 01:02:24

Diagnostics Check		OK	00:55
Air Blank		0.000	00:55
Control Test		0.081	00:56
Air Blank		0.000	00:56
Subject Sample	# 1	0.205	00:57
Breath Volume		1.507	00:57
Air Blank		0.000	00:58
Air Blank		0.000	01:00
Subject Sample	# 2	0.198	01:00
Breath Volume		1.722	01:00
Air Blank		0.000	01:01
Control Test		0.079	01:01
Air Blank		0.000	01:02
Diagnostics Check		OK	01:02

Data Download Date/Time: 07/06/2016 10:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,DANNY,W
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: GRAY,DANNY
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 06/08/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: GARCIA-TABOADA

DATE/TIME OF TEST: 07/03/2016 02:03:58

Diagnostics Check	OK	01:57
Air Blank	0.000	01:58
Control Test	0.079	01:58
Air Blank	0.000	01:58
Subject Sample # 1	0.105	01:59
Breath Volume	1.792	01:59
Air Blank	0.000	01:59
Air Blank	0.000	02:01
Subject Sample # 2	0.105	02:02
Breath Volume	2.011	02:02
Air Blank	0.000	02:03
Control Test	0.080	02:03
Air Blank	0.000	02:03
Diagnostics Check	OK	02:03

Data Download Date/Time: 07/06/2016 10:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FESSEL,DANIEL,C
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: FESSEL,DANIEL
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 06/08/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:19

SUBJECT'S LAST NAME: ORTIZ BERNAL

DATE/TIME OF TEST: 07/03/2016 03:47:29

Diagnostics Check	OK	03:39
Air Blank	0.000	03:40
Control Test	0.080	03:40
Air Blank	0.000	03:41
Subject Sample # 1	0.172	03:42
Breath Volume	1.378	03:42
Air Blank	0.000	03:43
Air Blank	0.000	03:45
Subject Sample # 2	0.163	03:45
Breath Volume	1.722	03:45
Air Blank	0.000	03:46
Control Test	0.079	03:46
Air Blank	0.000	03:47
Diagnostics Check	OK	03:47

Data Download Date/Time: 07/06/2016 10:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FILLAK, MICHAEL, O
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: ZARRILLO, NICHOLAS
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 06/08/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 19:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:31

SUBJECT'S LAST NAME: CARTER

DATE/TIME OF TEST: 07/03/2016 20:00:08

Diagnostics Check	OK	19:53
Air Blank	0.000	19:53
Control Test	0.081	19:54
Air Blank	0.000	19:54
Subject Sample # 1	0.160	19:55
Breath Volume	2.835	19:55
Air Blank	0.000	19:56
Air Blank	0.000	19:58
Subject Sample # 2	0.161	19:58
Breath Volume	3.378	19:58
Air Blank	0.000	19:59
Control Test	0.080	19:59
Air Blank	0.000	20:00
Diagnostics Check	OK	20:00

Data Download Date/Time: 07/06/2016 10:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAUSER,SGT,H
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: TACY,C
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 06/08/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:48

SUBJECT'S LAST NAME: WALTMAN

DATE/TIME OF TEST: 07/04/2016 22:15:17

Diagnostics Check	OK	22:08
Air Blank	0.000	22:08
Control Test	0.080	22:09
Air Blank	0.000	22:09
Subject Sample # 1	0.140	22:10
Breath Volume	2.386	22:10
Air Blank	0.000	22:11
Air Blank	0.000	22:12
Subject Sample # 2	0.140	22:13
Breath Volume	1.984	22:13
Air Blank	0.000	22:14
Control Test	0.080	22:14
Air Blank	0.000	22:15
Diagnostics Check	OK	22:15

Data Download Date/Time: 07/06/2016 10:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUSKEVICH,CURT,E
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: JOHNSON,S
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 06/08/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:08

SUBJECT'S LAST NAME: KULIKOWSKI

DATE/TIME OF TEST: 07/05/2016 02:44:52

Diagnostics Check	OK	02:38	
Air Blank	0.000	02:39	
Control Test	0.080	02:39	
Air Blank	0.000	02:39	
Subject Sample	# 1 0.234	02:40	
Breath Volume	2.738	02:40	
Air Blank	0.000	02:41	
Air Blank	0.000	02:42	
Subject Sample	# 2	02:43	Slope Not Met
Breath Volume	0.609	02:43	
Air Blank	0.000	02:43	
Control Test	0.080	02:44	
Air Blank	0.000	02:44	
Diagnostics Check	OK	02:44	

Data Download Date/Time: 07/06/2016 10:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUSKEVICH,CURT,E
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: JOHNSON,S
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 06/08/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:08

SUBJECT'S LAST NAME: KULIKOWSKI

DATE/TIME OF TEST: 07/05/2016 02:54:19

Diagnostics Check		OK	02:47
Air Blank		0.000	02:48
Control Test		0.079	02:48
Air Blank		0.000	02:49
Subject Sample	# 1	0.239	02:49
Breath Volume		2.984	02:49
Air Blank		0.000	02:50
Air Blank		0.000	02:52
Subject Sample	# 2	0.231	02:52
Breath Volume		3.257	02:52
Air Blank		0.000	02:53
Control Test		0.081	02:53
Air Blank		0.000	02:54
Diagnostics Check		OK	02:54

Data Download Date/Time: 07/06/2016 10:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAUSER,SGT,H
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: DENNIS,B
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 06/08/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 18:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:23

SUBJECT'S LAST NAME: DOLAN

DATE/TIME OF TEST: 07/05/2016 20:00:04

Diagnostics Check	OK	19:53
Air Blank	0.000	19:54
Control Test	0.080	19:54
Air Blank	0.000	19:54
Subject Sample # 1	0.000	19:55
Breath Volume	1.843	19:55
Air Blank	0.000	19:56
Air Blank	0.000	19:58
Subject Sample # 2	0.000	19:58
Breath Volume	2.277	19:58
Air Blank	0.000	19:59
Control Test	0.081	19:59
Air Blank	0.000	19:59
Diagnostics Check	OK	20:00

Data Download Date/Time: 07/06/2016 10:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,DANNY,W
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: GRAY,DANNY
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 07/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: EYERMAN

DATE/TIME OF TEST: 07/08/2016 05:06:42

Diagnostics Check	OK	05:00
Air Blank	0.000	05:01
Control Test	0.080	05:01
Air Blank	0.000	05:02
Subject Sample # 1	0.000	05:02
Breath Volume	2.847	05:02
Air Blank	0.000	05:02
Air Blank	0.000	05:04
Subject Sample # 2	0.000	05:05
Breath Volume	2.871	05:05
Air Blank	0.000	05:05
Control Test	0.081	05:06
Air Blank	0.000	05:06
Diagnostics Check	OK	05:06

Data Download Date/Time: 08/12/2016 10:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FESSEL,DANIEL,C
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: THOMAS,DANIEL
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 07/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:04

SUBJECT'S LAST NAME: YOUNG

DATE/TIME OF TEST: 07/09/2016 23:55:52

Diagnostics Check	OK	23:49
Air Blank	0.000	23:50
Control Test	0.079	23:50
Air Blank	0.000	23:50
Subject Sample # 1	0.153	23:51
Breath Volume	2.093	23:51
Air Blank	0.000	23:52
Air Blank	0.000	23:54
Subject Sample # 2	0.148	23:54
Breath Volume	1.347	23:54
Air Blank	0.000	23:54
Control Test	0.081	23:55
Air Blank	0.000	23:55
Diagnostics Check	OK	23:55

Data Download Date/Time: 08/12/2016 10:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUSKEVICH,CURT,E
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: THOMAS,D
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 07/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:57

SUBJECT'S LAST NAME: HUSTON

DATE/TIME OF TEST: 07/10/2016 04:25:42

Diagnostics Check	OK	04:19
Air Blank	0.000	04:19
Control Test	0.081	04:20
Air Blank	0.000	04:20
Subject Sample # 1	0.174	04:21
Breath Volume	4.144	04:21
Air Blank	0.000	04:21
Air Blank	0.000	04:23
Subject Sample # 2	0.178	04:24
Breath Volume	4.078	04:24
Air Blank	0.000	04:24
Control Test	0.080	04:25
Air Blank	0.000	04:25
Diagnostics Check	OK	04:25

Data Download Date/Time: 08/12/2016 10:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUSKEVICH,CURT,E
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: JOHNSON,S
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 07/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 07/10/2016 23:26:58

Diagnostics Check	OK	23:23	
Air Blank	0.000	23:23	
Control Test	0.080	23:23	
Air Blank	0.000	23:24	
Subject Sample	# 1	23:25	Subject Test Refused
Breath Volume	0.000	23:25	
Air Blank	0.000	23:26	
Control Test	0.080	23:26	
Air Blank	0.000	23:26	
Diagnostics Check	OK	23:26	

Data Download Date/Time: 08/12/2016 10:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CANNON,JAMES,J
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: TACY,CODY
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 07/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: PELLIS

DATE/TIME OF TEST: 07/14/2016 00:59:37

Diagnostics Check	OK	00:53
Air Blank	0.000	00:54
Control Test	0.080	00:54
Air Blank	0.000	00:54
Subject Sample # 1	0.104	00:55
Breath Volume	3.761	00:55
Air Blank	0.000	00:55
Air Blank	0.000	00:57
Subject Sample # 2	0.104	00:58
Breath Volume	3.210	00:58
Air Blank	0.000	00:58
Control Test	0.080	00:59
Air Blank	0.000	00:59
Diagnostics Check	OK	00:59

Data Download Date/Time: 08/12/2016 10:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FESSEL,DANIEL,C
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: THOMAS,DANIEL
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 07/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: STOUGHTON

DATE/TIME OF TEST: 07/14/2016 02:21:03

Diagnostics Check	OK	02:14
Air Blank	0.000	02:14
Control Test	0.080	02:14
Air Blank	0.000	02:15
Subject Sample # 1	0.187	02:15
Breath Volume	1.363	02:15
Air Blank	0.000	02:16
Air Blank	0.000	02:18
Subject Sample # 2	0.173	02:19
Breath Volume	1.273	02:19
Air Blank	0.000	02:20
Control Test	0.080	02:20
Air Blank	0.000	02:20
Diagnostics Check	OK	02:21

Data Download Date/Time: 08/12/2016 10:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARKLEY,JEFFREY,G
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MILLS,CHARLES
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 07/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: VALOE-RODRIGUEZ

DATE/TIME OF TEST: 07/16/2016 00:07:27

Diagnostics Check	OK	00:00
Air Blank	0.000	00:00
Control Test	0.081	00:01
Air Blank	0.000	00:01
Subject Sample # 1	0.098	00:02
Breath Volume	1.683	00:02
Air Blank	0.000	00:03
Air Blank	0.000	00:05
Subject Sample # 2	0.096	00:06
Breath Volume	1.867	00:06
Air Blank	0.000	00:06
Control Test	0.080	00:06
Air Blank	0.000	00:07
Diagnostics Check	OK	00:07

Data Download Date/Time: 08/12/2016 10:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARKLEY,JEFFREY,G
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: SATTERLEE,ANDREW
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 07/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:49

SUBJECT'S LAST NAME: FISCHER

DATE/TIME OF TEST: 07/16/2016 00:16:30

Diagnostics Check	OK	00:09
Air Blank	0.000	00:10
Control Test	0.080	00:10
Air Blank	0.000	00:11
Subject Sample # 1	0.024	00:12
Breath Volume	1.105	00:12
Air Blank	0.000	00:12
Air Blank	0.000	00:14
Subject Sample # 2	0.025	00:15
Breath Volume	1.574	00:15
Air Blank	0.000	00:15
Control Test	0.080	00:15
Air Blank	0.000	00:16
Diagnostics Check	OK	00:16

Data Download Date/Time: 08/12/2016 10:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,DANNY,W
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: GRAY,DANNY
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 07/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: RUBICCO

DATE/TIME OF TEST: 07/17/2016 03:18:21

Diagnostics Check	OK	03:12
Air Blank	0.000	03:12
Control Test	0.080	03:13
Air Blank	0.000	03:13
Subject Sample # 1	0.196	03:13
Breath Volume	2.695	03:13
Air Blank	0.000	03:14
Air Blank	0.000	03:16
Subject Sample # 2	0.192	03:16
Breath Volume	2.488	03:16
Air Blank	0.000	03:17
Control Test	0.079	03:17
Air Blank	0.000	03:18
Diagnostics Check	OK	03:18

Data Download Date/Time: 08/12/2016 10:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUSKEVICH,CURT,E
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: JOHNSON,S
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 07/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: LALLO

DATE/TIME OF TEST: 07/20/2016 03:45:44

Diagnostics Check	OK	03:42	
Air Blank	0.000	03:43	
Control Test	0.080	03:43	
Air Blank	0.000	03:44	
Subject Sample	# 1	03:44	Subject Test Refused
Breath Volume	0.000	03:44	
Air Blank	0.000	03:44	
Control Test	0.080	03:45	
Air Blank	0.000	03:45	
Diagnostics Check	OK	03:45	

Data Download Date/Time: 08/12/2016 10:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLS,CHARLES,S
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MILLS,CHARLES
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 07/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:52

SUBJECT'S LAST NAME: SOTO

DATE/TIME OF TEST: 07/23/2016 01:29:09

Diagnostics Check	OK	01:20
Air Blank	0.000	01:20
Control Test	0.080	01:21
Air Blank	0.000	01:21
Subject Sample # 1	0.177	01:22
Breath Volume	1.605	01:22
Air Blank	0.000	01:23
Air Blank	0.000	01:25
Subject Sample # 2	0.179	01:27
Breath Volume	2.578	01:27
Air Blank	0.000	01:28
Control Test	0.080	01:28
Air Blank	0.000	01:29
Diagnostics Check	OK	01:29

Data Download Date/Time: 08/12/2016 10:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLS,CHARLES,S
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: THOMAS,DANIEL
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 07/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: RAUSCH

DATE/TIME OF TEST: 07/23/2016 03:51:14

Diagnostics Check	OK	03:44	
Air Blank	0.000	03:44	
Control Test	0.080	03:45	
Air Blank	0.000	03:45	
Subject Sample # 1	0.174	03:46	
Breath Volume	1.707	03:46	
Air Blank	0.000	03:46	
Air Blank	0.000	03:48	
Subject Sample # 2	0.194	03:49	
Breath Volume	3.023	03:49	
Air Blank	0.000	03:50	
Control Test	0.079	03:50	
Air Blank		03:50	RFI Detect
Air Blank	0.000	03:51	

Data Download Date/Time: 08/12/2016 10:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLS,CHARLES,S
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: THOMAS,DANIEL
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 07/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: RAUSCH

DATE/TIME OF TEST: 07/23/2016 04:01:31

Diagnostics Check	OK	03:55
Air Blank	0.000	03:55
Control Test	0.080	03:55
Air Blank	0.000	03:56
Subject Sample # 1	0.190	03:56
Breath Volume	2.468	03:56
Air Blank	0.000	03:57
Air Blank	0.000	03:59
Subject Sample # 2	0.192	04:00
Breath Volume	3.191	04:00
Air Blank	0.000	04:00
Control Test	0.080	04:00
Air Blank	0.000	04:01
Diagnostics Check	OK	04:01

Data Download Date/Time: 08/12/2016 10:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CANNON,JAMES,J
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: JOHNSON,SARA
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 07/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:43

SUBJECT'S LAST NAME: RUEDA

DATE/TIME OF TEST: 07/23/2016 04:27:28

Diagnostics Check	OK	04:18
Air Blank	0.000	04:18
Control Test	0.080	04:19
Air Blank	0.000	04:19
Subject Sample # 1	0.098	04:22
Breath Volume	1.230	04:22
Air Blank	0.000	04:22
Air Blank	0.000	04:24
Subject Sample # 2	0.096	04:25
Breath Volume	1.296	04:25
Air Blank	0.000	04:26
Control Test	0.080	04:26
Air Blank	0.000	04:27
Diagnostics Check	OK	04:27

Data Download Date/Time: 08/12/2016 10:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUSKEVICH,CURT,E
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: JOHNSON,S
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 07/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:54

SUBJECT'S LAST NAME: WOLF

DATE/TIME OF TEST: 07/25/2016 04:20:11

Diagnostics Check	OK	04:16	
Air Blank	0.000	04:16	
Control Test	0.081	04:16	
Air Blank	0.000	04:17	
Subject Sample	# 1	04:18	Subject Test Refused
Breath Volume	0.000	04:18	
Air Blank	0.000	04:19	
Control Test	0.080	04:19	
Air Blank	0.000	04:20	
Diagnostics Check	OK	04:20	

Data Download Date/Time: 08/12/2016 10:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALLACE,JASON,A
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: WALLACE,JASON
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 07/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: MELENDEZ CRUZ

DATE/TIME OF TEST: 07/30/2016 03:38:17

Diagnostics Check	OK	03:30
Air Blank	0.000	03:31
Control Test	0.079	03:31
Air Blank	0.000	03:32
Subject Sample # 1	0.158	03:33
Breath Volume	1.781	03:33
Air Blank	0.000	03:34
Air Blank	0.000	03:36
Subject Sample # 2	0.157	03:36
Breath Volume	1.527	03:36
Air Blank	0.000	03:37
Control Test	0.078	03:37
Air Blank	0.000	03:38
Diagnostics Check	OK	03:38

Data Download Date/Time: 08/12/2016 10:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MINGS,DEP. G.,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: MINGS,DEP. G.
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 06/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A2
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:04

SUBJECT'S LAST NAME: VASQUEZ-QUISTIAN

DATE/TIME OF TEST: 07/03/2016 22:28:12

Diagnostics Check	OK	22:24	
Air Blank	0.000	22:25	
Control Test	0.078	22:25	
Air Blank	0.000	22:26	
Subject Sample	# 1	22:26	Subject Test Refused
Breath Volume	0.000	22:26	
Air Blank	0.000	22:27	
Control Test	0.080	22:27	
Air Blank	0.000	22:28	
Diagnostics Check	OK	22:28	

Data Download Date/Time: 07/29/2016 10:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000862

INSTRUMENT REGISTERED TO: BLOUNTSTOWN PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARSEY, WESLEY, H
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: HARSEY, WESLEY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 06/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG425202
DRY GAS EXPIRY DATE: 09/09/2016
VIOLATION CODE: DUI
ARREST TIME: 18:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:10

SUBJECT'S LAST NAME: MILES

DATE/TIME OF TEST: 07/10/2016 18:41:08

Diagnostics Check	OK	18:38	
Air Blank	0.000	18:38	
Control Test	0.080	18:39	
Air Blank	0.000	18:39	
Subject Sample	# 1	18:39	Subject Test Refused
Breath Volume	0.000	18:39	
Air Blank	0.000	18:40	
Control Test	0.079	18:40	
Air Blank	0.000	18:41	
Diagnostics Check	OK	18:41	

Data Download Date/Time: 08/02/2016 00:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000862

INSTRUMENT REGISTERED TO: BLOUNTSTOWN PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOLDEN,ANTHONY,E
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: GOLDEN,ANTHONY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 06/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG425202
DRY GAS EXPIRY DATE: 09/09/2016
VIOLATION CODE: DUI
ARREST TIME: 22:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:27

SUBJECT'S LAST NAME: MARTIN

DATE/TIME OF TEST: 07/31/2016 22:54:23

Diagnostics Check	OK	22:47
Air Blank	0.000	22:48
Control Test	0.080	22:48
Air Blank	0.000	22:49
Subject Sample # 1	0.186	22:49
Breath Volume	3.257	22:49
Air Blank	0.000	22:50
Air Blank	0.000	22:52
Subject Sample # 2	0.170	22:52
Breath Volume	1.679	22:52
Air Blank	0.000	22:53
Control Test	0.079	22:53
Air Blank	0.000	22:54
Diagnostics Check	OK	22:54

Data Download Date/Time: 08/02/2016 00:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORELL,JONAS,R
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: HEIN,PAUL
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 06/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1816818
DRY GAS EXPIRY DATE: 02/13/2018
VIOLATION CODE: BUI
ARREST TIME: 17:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:10

SUBJECT'S LAST NAME: HOLLIDAY

DATE/TIME OF TEST: 07/01/2016 18:34:10

Diagnostics Check	OK	18:31	
Air Blank	0.000	18:31	
Control Test	0.081	18:32	
Air Blank	0.000	18:32	
Subject Sample	# 1	18:32	Subject Test Refused
Breath Volume	0.000	18:32	
Air Blank	0.000	18:33	
Control Test	0.080	18:33	
Air Blank	0.000	18:34	
Diagnostics Check	OK	18:34	

Data Download Date/Time: 07/05/2016 07:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOORE,BRANDON,P
BREATH TEST OPERATOR AGENCY: Everglades Nat Park
ARRESTING OFFICER: MC NEIL,SEAN
ARRESTING OFFICER AGENCY: Everglades Nat Park
LAST AGENCY INSPECTION: 06/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1816818
DRY GAS EXPIRY DATE: 02/13/2018
VIOLATION CODE: BUI
ARREST TIME: 17:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:47

SUBJECT'S LAST NAME: YORK

DATE/TIME OF TEST: 07/02/2016 18:53:49

Diagnostics Check	OK	18:51	
Air Blank	0.000	18:51	
Control Test	0.081	18:51	
Air Blank	0.000	18:52	
Subject Sample	# 1	18:52	Subject Test Refused
Breath Volume	0.000	18:52	
Air Blank	0.000	18:52	
Control Test	0.080	18:53	
Air Blank	0.000	18:53	
Diagnostics Check	OK	18:53	

Data Download Date/Time: 07/05/2016 07:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORBETA,JORGE,
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: ORBETA,JORGE
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 06/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1816818
DRY GAS EXPIRY DATE: 02/13/2018
VIOLATION CODE: DUI
ARREST TIME: 22:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: ARGO

DATE/TIME OF TEST: 07/02/2016 23:48:42

Diagnostics Check	OK	23:37	
Air Blank	0.000	23:37	
Control Test	0.080	23:37	
Air Blank	0.000	23:38	
Subject Sample	# 1 0.182	23:41	Volume Not Met
Breath Volume	0.066	23:41	
Air Blank	0.000	23:42	
Air Blank	0.000	23:44	
Subject Sample	# 2 0.181	23:47	Volume Not Met
Breath Volume	0.117	23:47	
Air Blank	0.000	23:47	
Control Test	0.076	23:48	
Air Blank	0.000	23:48	
Diagnostics Check	OK	23:48	

Data Download Date/Time: 07/05/2016 07:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORBETA,JORGE,
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: ORBETA,JORGE
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 06/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1816818
DRY GAS EXPIRY DATE: 02/13/2018
VIOLATION CODE: DUI
ARREST TIME: 05:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:38

SUBJECT'S LAST NAME: CARRAZANA

DATE/TIME OF TEST: 07/03/2016 06:04:53

Diagnostics Check	OK	05:59	
Air Blank	0.000	06:00	
Control Test	0.080	06:00	
Air Blank	0.000	06:01	
Subject Sample	# 1	06:03	Subject Test Refused
Breath Volume	0.000	06:03	
Air Blank	0.000	06:03	
Control Test	0.080	06:04	
Air Blank	0.000	06:04	
Diagnostics Check	OK	06:04	

Data Download Date/Time: 07/05/2016 07:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOORE,BRANDON,P
BREATH TEST OPERATOR AGENCY: Everglades Nat Park
ARRESTING OFFICER: MOORE,BRANDON
ARRESTING OFFICER AGENCY: Everglades Nat Park
LAST AGENCY INSPECTION: 06/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1816818
DRY GAS EXPIRY DATE: 02/13/2018
VIOLATION CODE: BUI
ARREST TIME: 16:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:20

SUBJECT'S LAST NAME: REYES

DATE/TIME OF TEST: 07/03/2016 17:33:56

Diagnostics Check	OK	17:30	
Air Blank	0.000	17:31	
Control Test	0.080	17:31	
Air Blank	0.000	17:32	
Subject Sample	# 1	17:32	Subject Test Refused
Breath Volume	0.000	17:32	
Air Blank	0.000	17:33	
Control Test	0.080	17:33	
Air Blank	0.000	17:33	
Diagnostics Check	OK	17:33	

Data Download Date/Time: 07/05/2016 07:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORBETA,JORGE,
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: ORBETA,JORGE
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 06/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1816818
DRY GAS EXPIRY DATE: 02/13/2018
VIOLATION CODE: DUI
ARREST TIME: 22:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: AGIN

DATE/TIME OF TEST: 07/04/2016 00:06:39

Diagnostics Check	OK	00:00
Air Blank	0.000	00:00
Control Test	0.081	00:01
Air Blank	0.000	00:01
Subject Sample # 1	0.150	00:02
Breath Volume	1.816	00:02
Air Blank	0.000	00:02
Air Blank	0.000	00:04
Subject Sample # 2	0.148	00:05
Breath Volume	2.156	00:05
Air Blank	0.000	00:05
Control Test	0.079	00:06
Air Blank	0.000	00:06
Diagnostics Check	OK	00:06

Data Download Date/Time: 07/05/2016 07:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,GIL,A
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: GALVEZ,YUNIER
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 07/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1816818
DRY GAS EXPIRY DATE: 02/13/2018
VIOLATION CODE: DUI
ARREST TIME: 04:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:25

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 07/06/2016 05:52:56

Diagnostics Check	OK	05:50	
Air Blank	0.000	05:50	
Control Test	0.080	05:50	
Air Blank	0.000	05:51	
Subject Sample	# 1	05:51	Subject Test Refused
Breath Volume	0.000	05:51	
Air Blank	0.000	05:52	
Control Test	0.081	05:52	
Air Blank	0.000	05:52	
Diagnostics Check	OK	05:52	

Data Download Date/Time: 08/17/2016 23:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SEAGO,SHEILA,Y
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: TEST,GARRETT
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 07/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1816818
DRY GAS EXPIRY DATE: 02/13/2018
VIOLATION CODE: DUI
ARREST TIME: 04:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:32

SUBJECT'S LAST NAME: JOYCE

DATE/TIME OF TEST: 07/11/2016 04:56:19

Diagnostics Check	OK	04:53	
Air Blank	0.000	04:53	
Control Test	0.080	04:54	
Air Blank	0.000	04:54	
Subject Sample	# 1	04:54	Subject Test Refused
Breath Volume	0.000	04:54	
Air Blank	0.000	04:55	
Control Test	0.080	04:55	
Air Blank	0.000	04:56	
Diagnostics Check	OK	04:56	

Data Download Date/Time: 08/17/2016 23:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARRILLO,JULIO,C
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: CARRILLO,JULIO
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 07/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1816818
DRY GAS EXPIRY DATE: 02/13/2018
VIOLATION CODE: DUI
ARREST TIME: 09:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:47

SUBJECT'S LAST NAME: JACOBS

DATE/TIME OF TEST: 07/13/2016 11:14:20

Diagnostics Check	OK	11:08
Air Blank	0.000	11:08
Control Test	0.080	11:09
Air Blank	0.000	11:09
Subject Sample # 1	0.000	11:10
Breath Volume	2.542	11:10
Air Blank	0.000	11:10
Air Blank	0.000	11:12
Subject Sample # 2	0.000	11:12
Breath Volume	1.386	11:12
Air Blank	0.000	11:13
Control Test	0.081	11:13
Air Blank	0.000	11:14
Diagnostics Check	OK	11:14

Data Download Date/Time: 08/17/2016 23:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORBETA,JORGE,
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: ORBETA,JORGE
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 07/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1816818
DRY GAS EXPIRY DATE: 02/13/2018
VIOLATION CODE: DUI
ARREST TIME: 02:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: RUBINO

DATE/TIME OF TEST: 07/14/2016 03:22:33

Diagnostics Check		OK	03:10	
Air Blank		0.000	03:11	
Control Test		0.081	03:11	
Air Blank		0.000	03:11	
Subject Sample	# 1	0.010	03:14	Volume Not Met
Breath Volume		0.292	03:14	
Air Blank		0.000	03:15	
Air Blank		0.000	03:17	
Subject Sample	# 2	0.009	03:17	
Breath Volume		1.343	03:17	
Air Blank		0.000	03:18	
Air Blank		0.000	03:20	
Subject Sample	# 3	0.007	03:21	
Breath Volume		2.203	03:21	
Air Blank		0.000	03:21	
Control Test		0.079	03:21	
Air Blank		0.000	03:22	
Diagnostics Check		OK	03:22	

Data Download Date/Time: 08/17/2016 23:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRADY,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: ROCHA,PAUL
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 07/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1816818
DRY GAS EXPIRY DATE: 02/13/2018
VIOLATION CODE: DUI
ARREST TIME: 22:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:41

SUBJECT'S LAST NAME: ATCHLEY

DATE/TIME OF TEST: 07/15/2016 23:03:12

Diagnostics Check	OK	23:02	
Air Blank	0.000	23:02	
Control Test	0.000	23:02	RFI Detect
Air Blank	0.000	23:03	

Data Download Date/Time: 08/17/2016 23:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRADY,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: ROCHA,PAUL
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 07/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1816818
DRY GAS EXPIRY DATE: 02/13/2018
VIOLATION CODE: DUI
ARREST TIME: 22:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:41

SUBJECT'S LAST NAME: ATCHLEY

DATE/TIME OF TEST: 07/15/2016 23:12:19

Diagnostics Check	OK	23:06
Air Blank	0.000	23:06
Control Test	0.081	23:06
Air Blank	0.000	23:07
Subject Sample # 1	0.000	23:07
Breath Volume	3.218	23:07
Air Blank	0.000	23:08
Air Blank	0.000	23:10
Subject Sample # 2	0.000	23:10
Breath Volume	3.285	23:10
Air Blank	0.000	23:11
Control Test	0.080	23:11
Air Blank	0.000	23:12
Diagnostics Check	OK	23:12

Data Download Date/Time: 08/17/2016 23:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REINOSO,PEDRO,
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: REINOSO,PEDRO
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 07/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1816818
DRY GAS EXPIRY DATE: 02/13/2018
VIOLATION CODE: DUI
ARREST TIME: 02:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: CARTER

DATE/TIME OF TEST: 07/20/2016 03:12:29

Diagnostics Check	OK	03:08	
Air Blank	0.000	03:08	
Control Test	0.080	03:09	
Air Blank	0.000	03:09	
Subject Sample	# 1	03:10	Subject Test Refused
Breath Volume	0.972	03:10	
Air Blank	0.000	03:11	
Control Test	0.080	03:11	
Air Blank		03:12	RFI Detect
Air Blank	0.000	03:12	

Data Download Date/Time: 08/17/2016 23:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORBETA,JORGE,
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: ORBETA,JORGE
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 07/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1816818
DRY GAS EXPIRY DATE: 02/13/2018
VIOLATION CODE: DUI
ARREST TIME: 01:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:52

SUBJECT'S LAST NAME: PANSE

DATE/TIME OF TEST: 07/21/2016 02:15:55

Diagnostics Check	OK	02:12	
Air Blank	0.000	02:13	
Control Test	0.080	02:13	
Air Blank	0.000	02:14	
Subject Sample	# 1	02:14	Subject Test Refused
Breath Volume	0.000	02:14	
Air Blank	0.000	02:15	
Control Test	0.081	02:15	
Air Blank	0.000	02:15	
Diagnostics Check	OK	02:15	

Data Download Date/Time: 08/17/2016 23:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORBETA,JORGE,
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: ORBETA,JORGE
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 07/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1816818
DRY GAS EXPIRY DATE: 02/13/2018
VIOLATION CODE: DUI
ARREST TIME: 23:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:07

SUBJECT'S LAST NAME: AGRAFOJOS

DATE/TIME OF TEST: 07/24/2016 23:32:18

Diagnostics Check	OK	23:27	
Air Blank	0.000	23:28	
Control Test	0.080	23:28	
Air Blank	0.000	23:29	
Subject Sample	# 1	23:30	Subject Test Refused
Breath Volume	0.000	23:30	
Air Blank	0.000	23:31	
Control Test	0.081	23:31	
Air Blank	0.000	23:32	
Diagnostics Check	OK	23:32	

Data Download Date/Time: 08/17/2016 23:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KLEIS,KEVIN,W
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: KLEIS,KEVIN
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 07/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1816818
DRY GAS EXPIRY DATE: 02/13/2018
VIOLATION CODE: DUI
ARREST TIME: 04:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:00

SUBJECT'S LAST NAME: ZAMORA

DATE/TIME OF TEST: 07/27/2016 05:39:18

Diagnostics Check		OK	05:26	
Air Blank		0.000	05:26	
Control Test		0.081	05:26	
Air Blank		0.000	05:27	
Subject Sample	# 1	0.192	05:29	No .020 agreement
Breath Volume		1.421	05:29	
Air Blank		0.000	05:29	
Air Blank		0.000	05:31	
Subject Sample	# 2	0.177	05:34	Volume Not Met
Breath Volume		0.785	05:34	
Air Blank		0.000	05:35	
Air Blank		0.000	05:37	
Subject Sample	# 3	0.167	05:37	No .020 agreement
Breath Volume		1.457	05:37	
Air Blank		0.000	05:38	
Control Test		0.079	05:38	
Air Blank		0.000	05:39	
Diagnostics Check		OK	05:39	

Data Download Date/Time: 08/17/2016 23:28