Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: DAIL,EDD
ARRESTING OFFICER AGENCY: Oviedo PD
LAST AGENCY INSPECTION: 08/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 22:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:12

SUBJECT'S LAST NAME: DORSEY

DATE/TIME OF TEST: 09/02/2016 23:41:02

Diagnostics Check		OK	23:34
Air Blank		0.000	23:34
Control Test		0.080	23:35
Air Blank		0.000	23:35
Subject Sample	# 1	0.000	23:36
Breath Volume		2.531	23:36
Air Blank		0.000	23:37
Air Blank		0.000	23:39
Subject Sample	# 2	0.000	23:39
Breath Volume		1.570	23:39
Air Blank		0.000	23:40
Control Test		0.081	23:40
Air Blank		0.000	23:40
Diagnostics Check		OK	23:41

Data Download Date/Time: 09/30/2016 22:02

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: COLBURN,MEGAN
ARRESTING OFFICER AGENCY: Casselberry PD
LAST AGENCY INSPECTION: 08/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 01:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:43

SUBJECT'S LAST NAME: ANGIONE

DATE/TIME OF TEST: 09/03/2016 02:10:11

Diagnostics Check		OK	02:04
Air Blank		0.000	02:04
Control Test		0.081	02:04
Air Blank		0.000	02:05
Subject Sample	# 1	0.130	02:05
Breath Volume		2.359	02:05
Air Blank		0.000	02:06
Air Blank		0.000	02:08
Subject Sample	# 2	0.116	02:08
Breath Volume		1.726	02:08
Air Blank		0.000	02:09
Control Test		0.080	02:09
Air Blank		0.000	02:10
Diagnostics Check		OK	02:10

Data Download Date/Time: 09/30/2016 22:02

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: IVANOV,IVAN
ARRESTING OFFICER AGENCY: Sanford PD
LAST AGENCY INSPECTION: 08/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 20:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:35

SUBJECT'S LAST NAME: ALFARO-VELAQUEZ

DATE/TIME OF TEST: 09/03/2016 22:08:44

Diagnostics Check		OK	22:01
Air Blank		0.000	22:02
Control Test		0.082	22:02
Air Blank		0.000	22:03
Subject Sample	# 1	0.053	22:04
Breath Volume		2.246	22:04
Air Blank		0.000	22:04
Air Blank		0.000	22:06
Subject Sample	# 2	0.051	22:07
Breath Volume		1.429	22:07
Air Blank		0.000	22:07
Control Test		0.081	22:08
Air Blank		0.000	22:08
Diagnostics Check		OK	22:08

Data Download Date/Time: 09/30/2016 22:02

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: FRENCH,CHRISTOPHER
ARRESTING OFFICER AGENCY: Sanford PD
LAST AGENCY INSPECTION: 08/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 02:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:03

SUBJECT'S LAST NAME: MCCULLY

DATE/TIME OF TEST: 09/04/2016 03:31:39

Diagnostics Check		OK	03:25
Air Blank		0.000	03:25
Control Test		0.082	03:25
Air Blank		0.000	03:26
Subject Sample	# 1	0.206	03:27
Breath Volume		2.839	03:27
Air Blank		0.000	03:27
Air Blank		0.000	03:29
Subject Sample	# 2	0.213	03:30
Breath Volume		2.292	03:30
Air Blank		0.000	03:30
Control Test		0.080	03:31
Air Blank		0.000	03:31
Diagnostics Check		OK	03:31

Data Download Date/Time: 09/30/2016 22:02

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,RAY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: TOZZI,SHAUN
ARRESTING OFFICER AGENCY: Altamonte Springs PD
LAST AGENCY INSPECTION: 08/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 02:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: LIDDELL

DATE/TIME OF TEST: 09/07/2016 03:16:29

Diagnostics Check		OK	03:10
Air Blank		0.000	03:10
Control Test		0.082	03:10
Air Blank		0.000	03:11
Subject Sample	# 1	0.047	03:11
Breath Volume		1.957	03:11
Air Blank		0.000	03:12
Air Blank		0.000	03:14
Subject Sample	# 2	0.047	03:15
Breath Volume		2.296	03:15
Air Blank		0.000	03:15
Control Test		0.081	03:15
Air Blank		0.000	03:16
Diagnostics Check		OK	03:16

Data Download Date/Time: 09/30/2016 22:02

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,RAY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: TOZZI,SHAUN
ARRESTING OFFICER AGENCY: Altamonte Springs PD
LAST AGENCY INSPECTION: 08/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 23:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:37

SUBJECT'S LAST NAME: COOPER

DATE/TIME OF TEST: 09/08/2016 01:05:00

Diagnostics Check		OK	00:58
Air Blank		0.000	00:59
Control Test		0.081	00:59
Air Blank		0.000	01:00
Subject Sample	# 1	0.078	01:00
Breath Volume		1.875	01:00
Air Blank		0.000	01:01
Air Blank		0.000	01:03
Subject Sample	# 2	0.074	01:03
Breath Volume		1.886	01:03
Air Blank		0.000	01:04
Control Test		0.082	01:04
Air Blank		0.000	01:04
Diagnostics Check		OK	01:05

Data Download Date/Time: 09/30/2016 22:02

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,RAY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: REHDER,PATRICK
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 08/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 03:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:32

SUBJECT'S LAST NAME: BAEZ

DATE/TIME OF TEST: 09/08/2016 04:08:05

Diagnostics Check		OK	04:01
Air Blank		0.000	04:02
Control Test		0.081	04:02
Air Blank		0.000	04:02
Subject Sample	# 1	0.048	04:03
Breath Volume		1.234	04:03
Air Blank		0.000	04:03
Air Blank		0.000	04:05
Subject Sample	# 2	0.053	04:06
Breath Volume		3.054	04:06
Air Blank		0.000	04:07
Control Test		0.081	04:07
Air Blank		0.000	04:07
Diagnostics Check		OK	04:08

Data Download Date/Time: 09/30/2016 22:02

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: SIMPSON,CARL
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 08/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 20:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:53

SUBJECT'S LAST NAME: OSGOOD

DATE/TIME OF TEST: 09/08/2016 22:31:08

Diagnostics Check		OK	22:24
Air Blank		0.000	22:24
Control Test		0.081	22:25
Air Blank		0.000	22:25
Subject Sample	# 1	0.160	22:26
Breath Volume		1.226	22:26
Air Blank		0.000	22:27
Air Blank		0.000	22:28
Subject Sample	# 2	0.156	22:29
Breath Volume		1.800	22:29
Air Blank		0.000	22:30
Control Test		0.081	22:30
Air Blank		0.000	22:31
Diagnostics Check		OK	22:31

Data Download Date/Time: 09/30/2016 22:02

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: ANDREWS,GARY
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 08/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 03:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:40

SUBJECT'S LAST NAME: PRICE

DATE/TIME OF TEST: 09/09/2016 05:07:32

Diagnostics Check		OK	05:00
Air Blank		0.000	05:01
Control Test		0.081	05:01
Air Blank		0.000	05:02
Subject Sample	# 1	0.174	05:02
Breath Volume		5.355	05:02
Air Blank		0.000	05:03
Air Blank		0.000	05:05
Subject Sample	# 2	0.168	05:06
Breath Volume		5.007	05:06
Air Blank		0.000	05:06
Control Test		0.081	05:06
Air Blank		0.000	05:07
Diagnostics Check		OK	05:07

Data Download Date/Time: 09/30/2016 22:02

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: ZABALA,JOSE
ARRESTING OFFICER AGENCY: Oviedo PD
LAST AGENCY INSPECTION: 08/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 03:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:11

SUBJECT'S LAST NAME: CARTER

DATE/TIME OF TEST: 09/10/2016 04:38:25

Diagnostics Check		OK	04:32
Air Blank		0.000	04:32
Control Test		0.081	04:32
Air Blank		0.000	04:33
Subject Sample	# 1	0.216	04:34
Breath Volume		2.050	04:34
Air Blank		0.000	04:34
Air Blank		0.000	04:36
Subject Sample	# 2	0.224	04:36
Breath Volume		2.304	04:36
Air Blank		0.000	04:37
Control Test		0.081	04:37
Air Blank		0.000	04:38
Diagnostics Check		OK	04:38

Data Download Date/Time: 09/30/2016 22:02

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: ARGENTINE,BRITTANY
ARRESTING OFFICER AGENCY: Oviedo PD
LAST AGENCY INSPECTION: 08/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 04:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:56

SUBJECT'S LAST NAME: LEE

DATE/TIME OF TEST: 09/10/2016 05:22:51

Diagnostics Check		OK	05:16
Air Blank		0.000	05:17
Control Test		0.082	05:17
Air Blank		0.000	05:18
Subject Sample	# 1	0.185	05:18
Breath Volume		2.433	05:18
Air Blank		0.000	05:18
Air Blank		0.000	05:20
Subject Sample	# 2	0.185	05:21
Breath Volume		2.601	05:21
Air Blank		0.000	05:21
Control Test		0.081	05:22
Air Blank		0.000	05:22
Diagnostics Check		OK	05:22

Data Download Date/Time: 09/30/2016 22:02

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: RABORN,CHAD
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 08/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: ZERO TOL
ARREST TIME: 21:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:42

SUBJECT'S LAST NAME: MEJIA

DATE/TIME OF TEST: 09/10/2016 23:10:42

Diagnostics Check		OK	23:03
Air Blank		0.000	23:04
Control Test		0.082	23:04
Air Blank		0.000	23:05
Subject Sample	# 1	0.068	23:06
Breath Volume		1.273	23:06
Air Blank		0.000	23:06
Air Blank		0.000	23:08
Subject Sample	# 2	0.068	23:09
Breath Volume		1.972	23:09
Air Blank		0.000	23:09
Control Test		0.082	23:10
Air Blank		0.000	23:10
Diagnostics Check		OK	23:10

Data Download Date/Time: 09/30/2016 22:02

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,RAY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: LILJA,LORRI
ARRESTING OFFICER AGENCY: Lake Mary PD
LAST AGENCY INSPECTION: 08/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 02:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: STEIDLE

DATE/TIME OF TEST: 09/12/2016 03:22:24

Diagnostics Check		OK	03:16
Air Blank		0.000	03:16
Control Test		0.081	03:16
Air Blank		0.000	03:17
Subject Sample	# 1	0.217	03:18
Breath Volume		1.667	03:18
Air Blank		0.000	03:18
Air Blank		0.000	03:20
Subject Sample	# 2	0.210	03:20
Breath Volume		1.175	03:20
Air Blank		0.000	03:21
Control Test		0.081	03:21
Air Blank		0.000	03:22
Diagnostics Check		OK	03:22

Data Download Date/Time: 09/30/2016 22:02

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,RAY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: PASTRANA,RAYMOND
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 08/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 21:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: COMPTON

DATE/TIME OF TEST: 09/14/2016 23:11:42

Diagnostics Check		OK	23:05
Air Blank		0.000	23:06
Control Test		0.080	23:06
Air Blank		0.000	23:06
Subject Sample	# 1	0.091	23:07
Breath Volume		2.953	23:07
Air Blank		0.000	23:07
Air Blank		0.000	23:09
Subject Sample	# 2	0.090	23:10
Breath Volume		3.300	23:10
Air Blank		0.000	23:10
Control Test		0.081	23:11
Air Blank		0.000	23:11
Diagnostics Check		OK	23:11

Data Download Date/Time: 09/30/2016 22:02

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: BRONGEL,MATTHEW
ARRESTING OFFICER AGENCY: Oviedo PD
LAST AGENCY INSPECTION: 08/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 19:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:30

SUBJECT'S LAST NAME: ROMERO

DATE/TIME OF TEST: 09/15/2016 21:45:49

Diagnostics Check		OK	21:36
Air Blank		0.000	21:36
Control Test		0.081	21:37
Air Blank		0.000	21:37
Subject Sample	# 1	0.154	21:38
Breath Volume		1.121	21:38
Air Blank		0.000	21:38
Air Blank		0.000	21:40
Subject Sample	# 2	0.186	21:41
Breath Volume		3.488	21:41
Air Blank		0.000	21:41
Air Blank		0.000	21:43
Subject Sample	# 3	0.189	21:44
Breath Volume		3.554	21:44
Air Blank		0.000	21:44
Control Test		0.080	21:45
Air Blank		0.000	21:45
Diagnostics Check		OK	21:45

Data Download Date/Time: 09/30/2016 22:02

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: TOZZI,SHAUN
ARRESTING OFFICER AGENCY: Altamonte Springs PD
LAST AGENCY INSPECTION: 08/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 22:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:34

SUBJECT'S LAST NAME: FREI

DATE/TIME OF TEST: 09/17/2016 00:05:42

Diagnostics Check		OK	23:58
Air Blank		0.000	23:59
Control Test		0.081	23:59
Air Blank		0.000	00:00
Subject Sample	# 1	0.231	00:01
Breath Volume		1.937	00:01
Air Blank		0.000	00:01
Air Blank		0.000	00:03
Subject Sample	# 2	0.221	00:04
Breath Volume		1.308	00:04
Air Blank		0.000	00:04
Control Test		0.080	00:05
Air Blank		0.000	00:05
Diagnostics Check		OK	00:05

Data Download Date/Time: 09/30/2016 22:02

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: STAPP,MICHAEL
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 08/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 00:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:51

SUBJECT'S LAST NAME: WEEKS

DATE/TIME OF TEST: 09/17/2016 01:21:15

Diagnostics Check		OK	01:14
Air Blank		0.000	01:15
Control Test		0.081	01:15
Air Blank		0.000	01:15
Subject Sample	# 1	0.161	01:16
Breath Volume		2.992	01:16
Air Blank		0.000	01:17
Air Blank		0.000	01:19
Subject Sample	# 2	0.160	01:19
Breath Volume		2.140	01:19
Air Blank		0.000	01:20
Control Test		0.080	01:20
Air Blank		0.000	01:21
Diagnostics Check		OK	01:21

Data Download Date/Time: 09/30/2016 22:02

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: SPENCER,JOHN
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 08/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 02:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:19

SUBJECT'S LAST NAME: ELMORE

DATE/TIME OF TEST: 09/17/2016 03:46:12

Diagnostics Check		OK	03:40
Air Blank		0.000	03:40
Control Test		0.081	03:40
Air Blank		0.000	03:41
Subject Sample	# 1	0.113	03:41
Breath Volume		1.839	03:41
Air Blank		0.000	03:42
Air Blank		0.000	03:44
Subject Sample	# 2	0.111	03:44
Breath Volume		1.386	03:44
Air Blank		0.000	03:45
Control Test		0.081	03:45
Air Blank		0.000	03:46
Diagnostics Check		OK	03:46

Data Download Date/Time: 09/30/2016 22:02

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: REHDER,PATRICK
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 08/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 04:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:29

SUBJECT'S LAST NAME: HALL

DATE/TIME OF TEST: 09/17/2016 04:55:49

Diagnostics Check		OK	04:49
Air Blank		0.000	04:50
Control Test		0.080	04:50
Air Blank		0.000	04:50
Subject Sample	# 1	0.096	04:51
Breath Volume		2.238	04:51
Air Blank		0.000	04:51
Air Blank		0.000	04:53
Subject Sample	# 2	0.089	04:54
Breath Volume		1.507	04:54
Air Blank		0.000	04:54
Control Test		0.081	04:55
Air Blank		0.000	04:55
Diagnostics Check		OK	04:55

Data Download Date/Time: 09/30/2016 22:03

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: SHELDON,SPENCER
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 08/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 21:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:56

SUBJECT'S LAST NAME: IRIZARRY-MUNOZ

DATE/TIME OF TEST: 09/18/2016 02:32:45

Diagnostics Check		OK	02:26
Air Blank		0.000	02:27
Control Test		0.080	02:27
Air Blank		0.000	02:27
Subject Sample	# 1	0.149	02:28
Breath Volume		1.628	02:28
Air Blank		0.000	02:28
Air Blank		0.000	02:30
Subject Sample	# 2	0.150	02:31
Breath Volume		2.531	02:31
Air Blank		0.000	02:31
Control Test		0.081	02:32
Air Blank		0.000	02:32
Diagnostics Check		OK	02:32

Data Download Date/Time: 09/30/2016 22:03

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,RAY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: COLBURN,MEGAN
ARRESTING OFFICER AGENCY: Casselberry PD
LAST AGENCY INSPECTION: 08/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 20:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:35

SUBJECT'S LAST NAME: BULCOCK

DATE/TIME OF TEST: 09/18/2016 22:02:28

Diagnostics Check		OK	21:55
Air Blank		0.000	21:56
Control Test		0.080	21:56
Air Blank		0.000	21:57
Subject Sample	# 1	0.257	21:58
Breath Volume		2.628	21:58
Air Blank		0.000	21:58
Air Blank		0.000	22:00
Subject Sample	# 2	0.260	22:00
Breath Volume		2.660	22:00
Air Blank		0.000	22:01
Control Test		0.081	22:01
Air Blank		0.000	22:02
Diagnostics Check		OK	22:02

Data Download Date/Time: 09/30/2016 22:03

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,RAY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: FLANAGAN,BRENT
ARRESTING OFFICER AGENCY: Sanford PD
LAST AGENCY INSPECTION: 08/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 01:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: STAFFORD

DATE/TIME OF TEST: 09/19/2016 03:00:38

Diagnostics Check		OK	02:54
Air Blank		0.000	02:55
Control Test		0.081	02:55
Air Blank		0.000	02:56
Subject Sample	# 1	0.000	02:56
Breath Volume		1.453	02:56
Air Blank		0.000	02:56
Air Blank		0.000	02:58
Subject Sample	# 2	0.000	02:59
Breath Volume		1.265	02:59
Air Blank		0.000	02:59
Control Test		0.080	03:00
Air Blank		0.000	03:00
Diagnostics Check		OK	03:00

Data Download Date/Time: 09/30/2016 22:03

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,RAY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: MAYFIELD,JOHN
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 08/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 02:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: GERMAIN

DATE/TIME OF TEST: 09/22/2016 03:40:45

Diagnostics Check		OK	03:34
Air Blank		0.000	03:35
Control Test		0.081	03:35
Air Blank		0.000	03:36
Subject Sample	# 1	0.000	03:36
Breath Volume		2.238	03:36
Air Blank		0.000	03:37
Air Blank		0.000	03:39
Subject Sample	# 2	0.000	03:39
Breath Volume		2.503	03:39
Air Blank		0.000	03:39
Control Test		0.081	03:40
Air Blank		0.000	03:40
Diagnostics Check		OK	03:40

Data Download Date/Time: 09/30/2016 22:03

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,RAY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: RAGBIR,SHODAN
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 08/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 04:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:10

SUBJECT'S LAST NAME: RAMOS

DATE/TIME OF TEST: 09/22/2016 05:41:44

Diagnostics Check		OK	05:35
Air Blank		0.000	05:36
Control Test		0.080	05:36
Air Blank		0.000	05:37
Subject Sample	# 1	0.000	05:37
Breath Volume		2.054	05:37
Air Blank		0.000	05:38
Air Blank		0.000	05:40
Subject Sample	# 2	0.000	05:40
Breath Volume		2.312	05:40
Air Blank		0.000	05:40
Control Test		0.081	05:41
Air Blank		0.000	05:41
Diagnostics Check		OK	05:41

Data Download Date/Time: 09/30/2016 22:03

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: REHDER,PATRICK
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 08/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 01:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:46

SUBJECT'S LAST NAME: WALBERT

DATE/TIME OF TEST: 09/23/2016 02:19:35

Diagnostics Check		OK	02:11
Air Blank		0.000	02:11
Control Test		0.081	02:11
Air Blank		0.000	02:12
Subject Sample	# 1	0.219	02:13
Breath Volume		1.593	02:13
Air Blank		0.000	02:14
Air Blank		0.000	02:16
Subject Sample	# 2	0.223	02:18
Breath Volume		2.554	02:18
Air Blank		0.000	02:18
Control Test		0.082	02:19
Air Blank		0.000	02:19
Diagnostics Check		OK	02:19

Data Download Date/Time: 09/30/2016 22:03

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: TOZZI,SHAUN
ARRESTING OFFICER AGENCY: Altamonte Springs PD
LAST AGENCY INSPECTION: 08/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 02:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:04

SUBJECT'S LAST NAME: BOUCHARD

DATE/TIME OF TEST: 09/24/2016 03:31:01

Diagnostics Check		OK	03:24
Air Blank		0.000	03:25
Control Test		0.081	03:25
Air Blank		0.000	03:26
Subject Sample	# 1	0.099	03:26
Breath Volume		2.382	03:26
Air Blank		0.000	03:27
Air Blank		0.000	03:29
Subject Sample	# 2	0.095	03:29
Breath Volume		2.308	03:29
Air Blank		0.000	03:30
Control Test		0.082	03:30
Air Blank		0.000	03:30
Diagnostics Check		OK	03:31

Data Download Date/Time: 09/30/2016 22:03

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,RAY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: PHILLIPS,ADAM
ARRESTING OFFICER AGENCY: Casselberry PD
LAST AGENCY INSPECTION: 08/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 03:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:33

SUBJECT'S LAST NAME: HELDRETH

DATE/TIME OF TEST: 09/25/2016 03:59:12

Diagnostics Check		OK	03:53
Air Blank		0.000	03:53
Control Test		0.080	03:54
Air Blank		0.000	03:54
Subject Sample	# 1	0.149	03:54
Breath Volume		1.382	03:54
Air Blank		0.000	03:55
Air Blank		0.000	03:57
Subject Sample	# 2	0.139	03:57
Breath Volume		1.589	03:57
Air Blank		0.000	03:58
Control Test		0.081	03:58
Air Blank		0.000	03:59
Diagnostics Check		OK	03:59

Data Download Date/Time: 09/30/2016 22:03

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,RAY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: SPENCER,JOHN
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 08/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 02:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: MIRANDA CARRETO

DATE/TIME OF TEST: 09/25/2016 04:56:04

Diagnostics Check		OK	04:49
Air Blank		0.000	04:49
Control Test		0.081	04:49
Air Blank		0.000	04:50
Subject Sample	# 1	0.185	04:51
Breath Volume		1.105	04:51
Air Blank		0.000	04:51
Air Blank		0.000	04:53
Subject Sample	# 2	0.190	04:54
Breath Volume		1.886	04:54
Air Blank		0.000	04:55
Control Test		0.081	04:55
Air Blank		0.000	04:55
Diagnostics Check		OK	04:56

Data Download Date/Time: 09/30/2016 22:03

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,RAY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: RUIZ,IVAR
ARRESTING OFFICER AGENCY: Longwood PD
LAST AGENCY INSPECTION: 08/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 22:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:13

SUBJECT'S LAST NAME: GOUBEA

DATE/TIME OF TEST: 09/26/2016 00:39:01

Diagnostics Check		OK	00:33
Air Blank		0.000	00:33
Control Test		0.081	00:33
Air Blank		0.000	00:34
Subject Sample	# 1	0.130	00:34
Breath Volume		3.222	00:34
Air Blank		0.000	00:35
Air Blank		0.000	00:37
Subject Sample	# 2	0.133	00:37
Breath Volume		2.816	00:37
Air Blank		0.000	00:38
Control Test		0.081	00:38
Air Blank		0.000	00:38
Diagnostics Check		OK	00:39

Data Download Date/Time: 09/30/2016 22:03

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,RAY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: BRONGEL,MATTHEW
ARRESTING OFFICER AGENCY: Oviedo PD
LAST AGENCY INSPECTION: 08/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 03:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: HAAS

DATE/TIME OF TEST: 09/29/2016 04:38:14

Diagnostics Check		OK	04:32
Air Blank		0.000	04:32
Control Test		0.082	04:32
Air Blank		0.000	04:33
Subject Sample	# 1	0.151	04:33
Breath Volume		1.988	04:33
Air Blank		0.000	04:34
Air Blank		0.000	04:36
Subject Sample	# 2	0.155	04:36
Breath Volume		1.722	04:36
Air Blank		0.000	04:37
Control Test		0.081	04:37
Air Blank		0.000	04:38
Diagnostics Check		OK	04:38

Data Download Date/Time: 09/30/2016 22:03

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000999

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,RAY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: FRENCH,CHRISTOPHER
ARRESTING OFFICER AGENCY: Sanford PD
LAST AGENCY INSPECTION: 08/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1787697
DRY GAS EXPIRY DATE: 01/08/2018
VIOLATION CODE: DUI
ARREST TIME: 01:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:37

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 09/05/2016 02:04:13

Diagnostics Check		OK	01:57
Air Blank		0.000	01:57
Control Test		0.079	01:58
Air Blank		0.000	01:58
Subject Sample	# 1	0.125	01:59
Breath Volume		1.902	01:59
Air Blank		0.000	02:00
Air Blank		0.000	02:02
Subject Sample	# 2	0.125	02:02
Breath Volume		1.324	02:02
Air Blank		0.000	02:03
Control Test		0.079	02:03
Air Blank		0.000	02:04
Diagnostics Check		OK	02:04

Data Download Date/Time: 09/30/2016 22:06

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000999

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,RAY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: GUY,STEVEN
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 08/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1787697
DRY GAS EXPIRY DATE: 01/08/2018
VIOLATION CODE: DUI
ARREST TIME: 01:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:18

SUBJECT'S LAST NAME: RUBART

DATE/TIME OF TEST: 09/06/2016 02:44:28

Diagnostics Check		OK	02:38
Air Blank		0.000	02:38
Control Test		0.078	02:39
Air Blank		0.000	02:39
Subject Sample	# 1	0.077	02:40
Breath Volume		3.257	02:40
Air Blank		0.000	02:40
Air Blank		0.000	02:42
Subject Sample	# 2	0.077	02:43
Breath Volume		3.066	02:43
Air Blank		0.000	02:43
Control Test		0.079	02:43
Air Blank		0.000	02:44
Diagnostics Check		OK	02:44

Data Download Date/Time: 09/30/2016 22:06

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000999

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,RAY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: GRAHAM,KEYLAN
ARRESTING OFFICER AGENCY: Winter Springs PD
LAST AGENCY INSPECTION: 08/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1787697
DRY GAS EXPIRY DATE: 01/08/2018
VIOLATION CODE: DUI
ARREST TIME: 01:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:13

SUBJECT'S LAST NAME: TAORMINA

DATE/TIME OF TEST: 09/06/2016 03:07:12

Diagnostics Check		OK	03:01
Air Blank		0.000	03:01
Control Test		0.079	03:01
Air Blank		0.000	03:02
Subject Sample	# 1	0.149	03:02
Breath Volume		2.503	03:02
Air Blank		0.000	03:03
Air Blank		0.000	03:05
Subject Sample	# 2	0.149	03:05
Breath Volume		2.625	03:05
Air Blank		0.000	03:06
Control Test		0.078	03:06
Air Blank		0.000	03:07
Diagnostics Check		OK	03:07

Data Download Date/Time: 09/30/2016 22:06

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000999

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,RAY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: REHDER,PATRICK
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 08/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1787697
DRY GAS EXPIRY DATE: 01/08/2018
VIOLATION CODE: DUI
ARREST TIME: 03:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:51

SUBJECT'S LAST NAME: FOSTER

DATE/TIME OF TEST: 09/06/2016 04:21:31

Diagnostics Check		OK	04:11
Air Blank		0.000	04:11
Control Test		0.078	04:12
Air Blank		0.000	04:12
Subject Sample	# 1	0.145	04:13
Breath Volume		2.898	04:13
Air Blank		0.000	04:14
Air Blank		0.000	04:16
Subject Sample	# 2	0.118	04:16
Breath Volume		1.144	04:16
Air Blank		0.000	04:17
Air Blank		0.000	04:19
Subject Sample	# 3	0.141	04:20
Breath Volume		4.605	04:20
Air Blank		0.000	04:20
Control Test		0.079	04:20
Air Blank		0.000	04:21
Diagnostics Check		OK	04:21

Data Download Date/Time: 09/30/2016 22:06

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000999

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,RAY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: FLANAGAN,BRENT
ARRESTING OFFICER AGENCY: Sanford PD
LAST AGENCY INSPECTION: 08/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1787697
DRY GAS EXPIRY DATE: 01/08/2018
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: STANLEY

DATE/TIME OF TEST: 09/13/2016 02:40:27

Diagnostics Check		OK	02:34
Air Blank		0.000	02:34
Control Test		0.080	02:35
Air Blank		0.000	02:35
Subject Sample	# 1	0.000	02:36
Breath Volume		3.304	02:36
Air Blank		0.000	02:36
Air Blank		0.000	02:38
Subject Sample	# 2	0.000	02:39
Breath Volume		3.757	02:39
Air Blank		0.000	02:39
Control Test		0.079	02:39
Air Blank		0.000	02:40
Diagnostics Check		OK	02:40

Data Download Date/Time: 09/30/2016 22:06

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000999

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,RAY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: CASTEEL,JASON
ARRESTING OFFICER AGENCY: Winter Springs PD
LAST AGENCY INSPECTION: 08/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1787697
DRY GAS EXPIRY DATE: 01/08/2018
VIOLATION CODE: DUI
ARREST TIME: 01:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: CONTEEN

DATE/TIME OF TEST: 09/20/2016 03:11:27

Diagnostics Check		OK	03:04
Air Blank		0.000	03:05
Control Test		0.079	03:05
Air Blank		0.000	03:06
Subject Sample	# 1	0.128	03:07
Breath Volume		2.968	03:07
Air Blank		0.000	03:07
Air Blank		0.000	03:09
Subject Sample	# 2	0.127	03:09
Breath Volume		2.902	03:09
Air Blank		0.000	03:10
Control Test		0.079	03:10
Air Blank		0.000	03:11
Diagnostics Check		OK	03:11

Data Download Date/Time: 09/30/2016 22:06

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000999

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,RAY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: SPENCER,JOHN
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 08/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1787697
DRY GAS EXPIRY DATE: 01/08/2018
VIOLATION CODE: DUI
ARREST TIME: 02:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: EISMANN

DATE/TIME OF TEST: 09/21/2016 04:12:17

Diagnostics Check		OK	04:05
Air Blank		0.000	04:06
Control Test		0.078	04:06
Air Blank		0.000	04:07
Subject Sample	# 1	0.158	04:07
Breath Volume		1.136	04:07
Air Blank		0.000	04:08
Air Blank		0.000	04:09
Subject Sample	# 2	0.161	04:10
Breath Volume		2.078	04:10
Air Blank		0.000	04:11
Control Test		0.079	04:11
Air Blank		0.000	04:12
Diagnostics Check		OK	04:12

Data Download Date/Time: 09/30/2016 22:06

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000999

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,RAY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: SIMMONS,MAQIA
ARRESTING OFFICER AGENCY: Altamonte Springs PD
LAST AGENCY INSPECTION: 08/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1787697
DRY GAS EXPIRY DATE: 01/08/2018
VIOLATION CODE: DUI
ARREST TIME: 01:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:13

SUBJECT'S LAST NAME: BRASWELL

DATE/TIME OF TEST: 09/27/2016 02:39:48

Diagnostics Check		OK	02:33
Air Blank		0.000	02:34
Control Test		0.079	02:34
Air Blank		0.000	02:34
Subject Sample	# 1	0.184	02:35
Breath Volume		1.957	02:35
Air Blank		0.000	02:36
Air Blank		0.000	02:37
Subject Sample	# 2	0.191	02:38
Breath Volume		2.218	02:38
Air Blank		0.000	02:38
Control Test		0.079	02:39
Air Blank		0.000	02:39
Diagnostics Check		OK	02:39

Data Download Date/Time: 09/30/2016 22:06

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000999

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: TOZZI,SHAUN
ARRESTING OFFICER AGENCY: Altamonte Springs PD
LAST AGENCY INSPECTION: 09/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1787697
DRY GAS EXPIRY DATE: 01/08/2018
VIOLATION CODE: DUI
ARREST TIME: 00:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:19

SUBJECT'S LAST NAME: CANNON

DATE/TIME OF TEST: 09/30/2016 01:46:08

Diagnostics Check		OK	01:39
Air Blank		0.000	01:39
Control Test		0.079	01:40
Air Blank		0.000	01:40
Subject Sample	# 1	0.126	01:41
Breath Volume		1.503	01:41
Air Blank		0.000	01:42
Air Blank		0.000	01:44
Subject Sample	# 2	0.130	01:44
Breath Volume		1.785	01:44
Air Blank		0.000	01:45
Control Test		0.079	01:45
Air Blank		0.000	01:46
Diagnostics Check		OK	01:46

Data Download Date/Time: 09/30/2016 22:06

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000999

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: UMANA,CLAUDIA
ARRESTING OFFICER AGENCY: Lake Mary PD
LAST AGENCY INSPECTION: 09/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1787697
DRY GAS EXPIRY DATE: 01/08/2018
VIOLATION CODE: DUI
ARREST TIME: 04:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:46

SUBJECT'S LAST NAME: CASANOVA-BOLANOS

DATE/TIME OF TEST: 09/30/2016 05:13:00

Diagnostics Check		OK	05:06
Air Blank		0.000	05:06
Control Test		0.080	05:07
Air Blank		0.000	05:07
Subject Sample	# 1	0.123	05:08
Breath Volume		3.242	05:08
Air Blank		0.000	05:09
Air Blank		0.000	05:10
Subject Sample	# 2	0.119	05:11
Breath Volume		1.777	05:11
Air Blank		0.000	05:12
Control Test		0.079	05:12
Air Blank		0.000	05:12
Diagnostics Check		OK	05:13

Data Download Date/Time: 09/30/2016 22:06

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001000

INSTRUMENT REGISTERED TO: COCOA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,JAMES,H
BREATH TEST OPERATOR AGENCY: Cocoa PD
ARRESTING OFFICER: RUSH,MATT
ARRESTING OFFICER AGENCY: Cocoa PD
LAST AGENCY INSPECTION: 08/02/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A2
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:12

SUBJECT'S LAST NAME: WERTS

DATE/TIME OF TEST: 09/17/2016 21:53:16

Diagnostics Check		OK	21:50	
Air Blank		0.000	21:50	
Control Test		0.080	21:51	
Air Blank		0.000	21:51	
Subject Sample	# 1		21:51	Subject Test Refused
Breath Volume		0.000	21:51	
Air Blank		0.000	21:52	
Control Test		0.081	21:52	
Air Blank		0.000	21:53	
Diagnostics Check		OK	21:53	

Data Download Date/Time: 09/28/2016 20:45

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROOKS,ROBBIE,M
BREATH TEST OPERATOR AGENCY: Satellite Beach PD
ARRESTING OFFICER: MCCRANN,SEAN
ARRESTING OFFICER AGENCY: Satellite Beach PD
LAST AGENCY INSPECTION: 08/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1878845
DRY GAS EXPIRY DATE: 06/22/2018
VIOLATION CODE: DUI
ARREST TIME: 00:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:58

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 09/03/2016 01:56:55

Diagnostics Check		OK	01:48	
Air Blank		0.000	01:48	
Control Test		0.079	01:49	
Air Blank		0.000	01:49	
Subject Sample	# 1		01:52	No Sample Provided
Breath Volume		0.000	01:52	
Air Blank		0.000	01:53	
Air Blank		0.000	01:55	
Subject Sample	# 2		01:56	Sequence Aborted
Breath Volume		0.000	01:56	
Air Blank		0.000	01:56	

Data Download Date/Time: 09/30/2016 00:33

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROOKS,ROBBIE,M
BREATH TEST OPERATOR AGENCY: Satellite Beach PD
ARRESTING OFFICER: MCCRANN,SEAN
ARRESTING OFFICER AGENCY: Satellite Beach PD
LAST AGENCY INSPECTION: 08/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1878845
DRY GAS EXPIRY DATE: 06/22/2018
VIOLATION CODE: DUI
ARREST TIME: 00:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 09/03/2016 02:16:38

Diagnostics Check		OK	02:10
Air Blank		0.000	02:10
Control Test		0.078	02:11
Air Blank		0.000	02:11
Subject Sample	# 1	0.027	02:12
Breath Volume		1.812	02:12
Air Blank		0.000	02:12
Air Blank		0.000	02:14
Subject Sample	# 2	0.028	02:15
Breath Volume		1.976	02:15
Air Blank		0.000	02:15
Control Test		0.079	02:16
Air Blank		0.000	02:16
Diagnostics Check		OK	02:16

Data Download Date/Time: 09/30/2016 00:33

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,ANTHONY,M
BREATH TEST OPERATOR AGENCY: Satellite Beach PD
ARRESTING OFFICER: HOLLAND,ANTHONY
ARRESTING OFFICER AGENCY: Satellite Beach PD
LAST AGENCY INSPECTION: 08/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1878845
DRY GAS EXPIRY DATE: 06/22/2018
VIOLATION CODE: DUI
ARREST TIME: 15:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:26

SUBJECT'S LAST NAME: CORD

DATE/TIME OF TEST: 09/04/2016 15:54:02

Diagnostics Check		OK	15:47
Air Blank		0.000	15:48
Control Test		0.078	15:48
Air Blank		0.000	15:49
Subject Sample	# 1	0.000	15:49
Breath Volume		1.382	15:49
Air Blank		0.000	15:50
Air Blank		0.000	15:52
Subject Sample	# 2	0.000	15:52
Breath Volume		1.679	15:52
Air Blank		0.000	15:53
Control Test		0.078	15:53
Air Blank		0.000	15:53
Diagnostics Check		OK	15:54

Data Download Date/Time: 09/30/2016 00:33

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRIEBELL,DON,R
BREATH TEST OPERATOR AGENCY: Satellite Beach PD
ARRESTING OFFICER: TRIEBELL,DON
ARRESTING OFFICER AGENCY: Satellite Beach PD
LAST AGENCY INSPECTION: 08/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1878845
DRY GAS EXPIRY DATE: 06/22/2018
VIOLATION CODE: DUI
ARREST TIME: 23:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:48

SUBJECT'S LAST NAME: GERSTAD

DATE/TIME OF TEST: 09/10/2016 00:21:58

Diagnostics Check		OK	00:15
Air Blank		0.000	00:15
Control Test		0.077	00:16
Air Blank		0.000	00:16
Subject Sample	# 1	0.179	00:17
Breath Volume		2.058	00:17
Air Blank		0.000	00:18
Air Blank		0.000	00:19
Subject Sample	# 2	0.178	00:20
Breath Volume		1.601	00:20
Air Blank		0.000	00:21
Control Test		0.078	00:21
Air Blank		0.000	00:21
Diagnostics Check		OK	00:21

Data Download Date/Time: 09/30/2016 00:33

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCGANN,JAMES,D
BREATH TEST OPERATOR AGENCY: Satellite Beach PD
ARRESTING OFFICER: MCGANN,JAMES
ARRESTING OFFICER AGENCY: Satellite Beach PD
LAST AGENCY INSPECTION: 08/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1878845
DRY GAS EXPIRY DATE: 06/22/2018
VIOLATION CODE: DUI
ARREST TIME: 14:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:24

SUBJECT'S LAST NAME: HANSEN

DATE/TIME OF TEST: 09/15/2016 16:05:26

Diagnostics Check		OK	15:53	
Air Blank		0.000	15:54	
Control Test		0.078	15:54	
Air Blank		0.000	15:54	
Subject Sample	# 1	0.202	15:58	Volume Not Met
Breath Volume		0.113	15:58	
Air Blank		0.000	15:58	
Air Blank		0.000	16:00	
Subject Sample	# 2	0.155	16:03	Volume Not Met
Breath Volume		0.617	16:03	
Air Blank		0.000	16:04	
Control Test		0.076	16:04	
Air Blank		0.000	16:05	
Diagnostics Check		OK	16:05	

Data Download Date/Time: 09/30/2016 00:33

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCGANN,JAMES,D
BREATH TEST OPERATOR AGENCY: Satellite Beach PD
ARRESTING OFFICER: MCGANN,JAMES
ARRESTING OFFICER AGENCY: Satellite Beach PD
LAST AGENCY INSPECTION: 08/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1878845
DRY GAS EXPIRY DATE: 06/22/2018
VIOLATION CODE: DUI
ARREST TIME: 14:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:24

SUBJECT'S LAST NAME: HANSEN

DATE/TIME OF TEST: 09/15/2016 16:15:46

Diagnostics Check		OK	16:07
Air Blank		0.000	16:08
Control Test		0.075	16:08
Air Blank		0.000	16:09
Subject Sample	# 1	0.191	16:09
Breath Volume		2.402	16:09
Air Blank		0.000	16:10
Air Blank		0.000	16:12
Subject Sample	# 2	0.197	16:14
Breath Volume		2.226	16:14
Air Blank		0.000	16:14
Control Test		0.075	16:15
Air Blank		0.000	16:15
Diagnostics Check		OK	16:15

Data Download Date/Time: 09/30/2016 00:33

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROOKS,ROBBIE,M
BREATH TEST OPERATOR AGENCY: Satellite Beach PD
ARRESTING OFFICER: MEEHAN,RYAN
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 08/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1878845
DRY GAS EXPIRY DATE: 06/22/2018
VIOLATION CODE: DUI
ARREST TIME: 19:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:38

SUBJECT'S LAST NAME: MASSERIA

DATE/TIME OF TEST: 09/17/2016 21:19:42

Diagnostics Check		OK	21:13
Air Blank		0.000	21:14
Control Test		0.079	21:14
Air Blank		0.000	21:14
Subject Sample	# 1	0.152	21:15
Breath Volume		2.277	21:15
Air Blank		0.000	21:15
Air Blank		0.000	21:17
Subject Sample	# 2	0.155	21:18
Breath Volume		2.246	21:18
Air Blank		0.000	21:18
Control Test		0.079	21:19
Air Blank		0.000	21:19
Diagnostics Check		OK	21:19

Data Download Date/Time: 09/30/2016 00:33

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001002

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTWRIGHT JR,GARY,W
BREATH TEST OPERATOR AGENCY: Hardee County SO
ARRESTING OFFICER: BRANDEBERRY,JEREMY
ARRESTING OFFICER AGENCY: Hardee County SO
LAST AGENCY INSPECTION: 08/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 03:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: MASCORRO

DATE/TIME OF TEST: 09/09/2016 03:46:51

Diagnostics Check		OK	03:40
Air Blank		0.000	03:40
Control Test		0.081	03:41
Air Blank		0.000	03:41
Subject Sample	# 1	0.120	03:42
Breath Volume		3.175	03:42
Air Blank		0.000	03:42
Air Blank		0.000	03:44
Subject Sample	# 2	0.119	03:45
Breath Volume		2.644	03:45
Air Blank		0.000	03:45
Control Test		0.081	03:46
Air Blank		0.000	03:46
Diagnostics Check		OK	03:46

Data Download Date/Time: 09/28/2016 23:52

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PANIAGUA,SGT M.,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ASHCOM,DEP T.
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 08/27/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG605301
DRY GAS EXPIRY DATE: 02/22/2018
VIOLATION CODE: DUI
ARREST TIME: 22:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: WALKER

DATE/TIME OF TEST: 09/02/2016 22:44:58

Diagnostics Check		OK	22:41	
Air Blank		0.000	22:42	
Control Test		0.077	22:42	
Air Blank		0.000	22:43	
Subject Sample	# 1		22:43	Subject Test Refused
Breath Volume		0.000	22:43	
Air Blank		0.000	22:43	
Control Test		0.078	22:44	
Air Blank		0.000	22:44	
Diagnostics Check		OK	22:44	

Data Download Date/Time: 09/29/2016 23:35

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PANIAGUA,SGT. M,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: PERSAUD,DEP C.
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 08/27/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG605301
DRY GAS EXPIRY DATE: 02/22/2018
VIOLATION CODE: DUI
ARREST TIME: 00:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:12

SUBJECT'S LAST NAME: MARSHALL

DATE/TIME OF TEST: 09/03/2016 00:38:47

Diagnostics Check		OK	00:32
Air Blank		0.000	00:32
Control Test		0.077	00:33
Air Blank		0.000	00:33
Subject Sample	# 1	0.205	00:34
Breath Volume		1.933	00:34
Air Blank		0.000	00:34
Air Blank		0.000	00:36
Subject Sample	# 2	0.222	00:37
Breath Volume		1.839	00:37
Air Blank		0.000	00:37
Control Test		0.079	00:38
Air Blank		0.000	00:38
Diagnostics Check		OK	00:38

Data Download Date/Time: 09/29/2016 23:35

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PANIAGUA,SGT M.,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KASELAK,DEP S
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 08/27/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG605301
DRY GAS EXPIRY DATE: 02/22/2018
VIOLATION CODE: DUI
ARREST TIME: 03:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: FEASTER

DATE/TIME OF TEST: 09/03/2016 04:09:01

Diagnostics Check		OK	04:05	
Air Blank		0.000	04:06	
Control Test		0.078	04:06	
Air Blank		0.000	04:07	
Subject Sample	# 1		04:07	Subject Test Refused
Breath Volume		0.000	04:07	
Air Blank		0.000	04:08	
Control Test		0.079	04:08	
Air Blank		0.000	04:08	
Diagnostics Check		OK	04:09	

Data Download Date/Time: 09/29/2016 23:35

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001006

INSTRUMENT REGISTERED TO: INDIALANTIC P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLSTINE,SCOTT,
BREATH TEST OPERATOR AGENCY: Indialantic PD
ARRESTING OFFICER: JENNINGS,JOHN
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 08/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A1
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:01

SUBJECT'S LAST NAME: COLACIOS

DATE/TIME OF TEST: 09/04/2016 04:04:28

Diagnostics Check		OK	03:57
Air Blank		0.000	03:57
Control Test		0.078	03:58
Air Blank		0.000	03:58
Subject Sample	# 1	0.196	03:59
Breath Volume		3.425	03:59
Air Blank		0.000	04:00
Air Blank		0.000	04:02
Subject Sample	# 2	0.208	04:02
Breath Volume		2.042	04:02
Air Blank		0.000	04:03
Control Test		0.077	04:03
Air Blank		0.000	04:04
Diagnostics Check		OK	04:04

Data Download Date/Time: 09/04/2016 13:17

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001006

INSTRUMENT REGISTERED TO: INDIALANTIC P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUTTEBO,KENT,N
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: GUTTEBO,KENT
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 09/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A1
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:34

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 09/23/2016 23:10:47

Diagnostics Check		OK	23:02
Air Blank		0.000	23:02
Control Test		0.078	23:02
Air Blank		0.000	23:03
Subject Sample	# 1	0.300	23:05
Breath Volume		1.437	23:05
Air Blank		0.000	23:05
Air Blank		0.000	23:07
Subject Sample	# 2	0.300	23:09
Breath Volume		1.664	23:09
Air Blank		0.000	23:09
Control Test		0.077	23:10
Air Blank		0.000	23:10
Diagnostics Check		OK	23:10

Data Download Date/Time: 10/23/2016 14:36

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TURNER,LATORIAN,T
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: TURNER,LATORIAN
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 08/19/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 390314
DRY GAS EXPIRY DATE: 04/19/2019
VIOLATION CODE: DUI
ARREST TIME: 21:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: CARWISE

DATE/TIME OF TEST: 09/04/2016 23:03:26

Diagnostics Check		OK	22:56
Air Blank		0.000	22:57
Control Test		0.082	22:57
Air Blank		0.000	22:57
Subject Sample	# 1	0.272	22:58
Breath Volume		2.652	22:58
Air Blank		0.000	22:59
Air Blank		0.000	23:01
Subject Sample	# 2	0.260	23:01
Breath Volume		1.171	23:01
Air Blank		0.000	23:02
Control Test		0.081	23:02
Air Blank		0.000	23:03
Diagnostics Check		OK	23:03

Data Download Date/Time: 09/17/2016 04:53

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAGER,KEVIN,M
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: HEPTINSTALL,D
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 08/19/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 390314
DRY GAS EXPIRY DATE: 04/19/2019
VIOLATION CODE: DUI
ARREST TIME: 12:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:00

SUBJECT'S LAST NAME: SADOWSKA

DATE/TIME OF TEST: 09/11/2016 13:25:52

Diagnostics Check		OK	13:20
Air Blank		0.000	13:20
Control Test		0.082	13:20
Air Blank		0.000	13:21
Subject Sample	# 1	0.000	13:21
Breath Volume		1.714	13:21
Air Blank		0.000	13:22
Air Blank		0.000	13:24
Subject Sample	# 2	0.000	13:24
Breath Volume		1.531	13:24
Air Blank		0.000	13:24
Control Test		0.079	13:25
Air Blank		0.000	13:25
Diagnostics Check		OK	13:25

Data Download Date/Time: 09/17/2016 04:53

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TURNER,LATORIAN,T
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: TURNER,LATORIAN
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 09/17/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 390314
DRY GAS EXPIRY DATE: 04/19/2019
VIOLATION CODE: DUI
ARREST TIME: 04:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:35

SUBJECT'S LAST NAME: LOUINE

DATE/TIME OF TEST: 09/23/2016 06:04:35

Diagnostics Check		OK	05:58
Air Blank		0.000	05:58
Control Test		0.082	05:58
Air Blank		0.000	05:59
Subject Sample	# 1	0.084	06:00
Breath Volume		1.988	06:00
Air Blank		0.000	06:00
Air Blank		0.000	06:02
Subject Sample	# 2	0.082	06:02
Breath Volume		2.949	06:02
Air Blank		0.000	06:03
Control Test		0.083	06:03
Air Blank		0.000	06:04
Diagnostics Check		OK	06:04

Data Download Date/Time: 10/20/2016 07:37

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001009

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEST,CHRISTINA,A
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: SMITH,DUANE
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 08/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 03:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:22

SUBJECT'S LAST NAME: SANTIAGO

DATE/TIME OF TEST: 09/04/2016 05:02:47

Diagnostics Check		OK	04:57	
Air Blank		0.000	04:57	
Control Test		0.079	04:57	
Air Blank		0.000	04:58	
Subject Sample	# 1		05:01	Subject Test Refused
Breath Volume		0.453	05:01	
Air Blank		0.000	05:01	
Control Test		0.080	05:02	
Air Blank		0.000	05:02	
Diagnostics Check		OK	05:02	

Data Download Date/Time: 09/25/2016 11:35

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001009

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIDGEON,ANDREW,W
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: PIDGEON,ANDREW
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 08/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 07:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:10

SUBJECT'S LAST NAME: GREBLICK

DATE/TIME OF TEST: 09/09/2016 08:38:16

Diagnostics Check		OK	08:32
Air Blank		0.000	08:32
Control Test		0.079	08:33
Air Blank		0.000	08:33
Subject Sample	# 1	0.000	08:34
Breath Volume		2.496	08:34
Air Blank		0.000	08:34
Air Blank		0.000	08:36
Subject Sample	# 2	0.000	08:36
Breath Volume		1.871	08:36
Air Blank		0.000	08:37
Control Test		0.078	08:37
Air Blank		0.000	08:38
Diagnostics Check		OK	08:38

Data Download Date/Time: 09/25/2016 11:35

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001009

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VURRO,MARC,E
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: HOLLCROFT,DEREK
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 08/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 00:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:18

SUBJECT'S LAST NAME: KLEIN

DATE/TIME OF TEST: 09/17/2016 01:59:41

Diagnostics Check		OK	01:55	
Air Blank		0.000	01:56	
Control Test		0.080	01:56	
Air Blank		0.000	01:57	
Subject Sample	# 1		01:58	Subject Test Refused
Breath Volume		0.000	01:58	
Air Blank		0.000	01:58	
Control Test		0.080	01:59	
Air Blank		0.000	01:59	
Diagnostics Check		OK	01:59	

Data Download Date/Time: 09/25/2016 11:35

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001009

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VURRO,MARCELLO,E
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: FOOKS,TYLER
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 08/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 01:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:28

SUBJECT'S LAST NAME: BULL

DATE/TIME OF TEST: 09/24/2016 02:03:40

Diagnostics Check		OK	01:56
Air Blank		0.000	01:57
Control Test		0.080	01:57
Air Blank		0.000	01:57
Subject Sample	# 1	0.215	01:58
Breath Volume		1.406	01:58
Air Blank		0.000	01:59
Air Blank		0.000	02:00
Subject Sample	# 2	0.203	02:02
Breath Volume		1.679	02:02
Air Blank		0.000	02:02
Control Test		0.079	02:03
Air Blank		0.000	02:03
Diagnostics Check		OK	02:03

Data Download Date/Time: 09/25/2016 11:35

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESCOBAR,ALEJANDRO,
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf
ARRESTING OFFICER: ESCOBAR,ALEJANDRO
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf
LAST AGENCY INSPECTION: 08/16/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1698962
DRY GAS EXPIRY DATE: 06/22/2017
VIOLATION CODE: DUI
ARREST TIME: 03:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:27

SUBJECT'S LAST NAME: PINE

DATE/TIME OF TEST: 09/08/2016 03:59:35

Diagnostics Check		OK	03:56	
Air Blank		0.000	03:56	
Control Test		0.077	03:57	
Air Blank		0.000	03:57	
Subject Sample	# 1		03:58	Subject Test Refused
Breath Volume		0.000	03:58	
Air Blank		0.000	03:58	
Control Test		0.077	03:59	
Air Blank		0.000	03:59	
Diagnostics Check		OK	03:59	

Data Download Date/Time: 09/20/2016 10:53

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESCOBAR,ALEJANDRO,
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf
ARRESTING OFFICER: ESCOBAR,ALEJANDRO
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf
LAST AGENCY INSPECTION: 08/16/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1698962
DRY GAS EXPIRY DATE: 06/22/2017
VIOLATION CODE: DUI
ARREST TIME: 01:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:09

SUBJECT'S LAST NAME: VILLATORO

DATE/TIME OF TEST: 09/11/2016 02:39:53

Diagnostics Check		OK	02:37	
Air Blank		0.000	02:37	
Control Test		0.078	02:37	
Air Blank		0.000	02:38	
Subject Sample	# 1		02:38	Subject Test Refused
Breath Volume		0.000	02:38	
Air Blank		0.000	02:39	
Control Test		0.077	02:39	
Air Blank		0.000	02:39	
Diagnostics Check		OK	02:39	

Data Download Date/Time: 09/20/2016 10:53

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESCOBAR,ALEJANDRO,
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf
ARRESTING OFFICER: ESCOBAR,ALEJANDRO
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf
LAST AGENCY INSPECTION: 08/16/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1698962
DRY GAS EXPIRY DATE: 06/22/2017
VIOLATION CODE: DUI
ARREST TIME: 02:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:13

SUBJECT'S LAST NAME: LEGER

DATE/TIME OF TEST: 09/13/2016 03:44:00

Diagnostics Check		OK	03:41	
Air Blank		0.000	03:41	
Control Test		0.077	03:41	
Air Blank		0.000	03:42	
Subject Sample	# 1		03:42	Subject Test Refused
Breath Volume		0.000	03:42	
Air Blank		0.000	03:43	
Control Test		0.078	03:43	
Air Blank		0.000	03:43	
Diagnostics Check		OK	03:44	

Data Download Date/Time: 09/20/2016 10:53

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESCOBAR,ALEJANDRO,
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf
ARRESTING OFFICER: ESCOBAR,ALEJANDRO
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf
LAST AGENCY INSPECTION: 09/20/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1698962
DRY GAS EXPIRY DATE: 06/22/2017
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:23

SUBJECT'S LAST NAME: BENTIVEGNA

DATE/TIME OF TEST: 09/21/2016 01:56:49

Diagnostics Check		OK	01:50
Air Blank		0.000	01:50
Control Test		0.077	01:51
Air Blank		0.000	01:51
Subject Sample	# 1	0.069	01:52
Breath Volume		1.699	01:52
Air Blank		0.000	01:52
Air Blank		0.000	01:54
Subject Sample	# 2	0.064	01:55
Breath Volume		1.191	01:55
Air Blank		0.000	01:55
Control Test		0.077	01:56
Air Blank		0.000	01:56
Diagnostics Check		OK	01:56

Data Download Date/Time: 10/24/2016 12:14

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESCOBAR,ALEJANDRO,
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf
ARRESTING OFFICER: ESCOBAR,ALEJANDRO
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf
LAST AGENCY INSPECTION: 09/20/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1698962
DRY GAS EXPIRY DATE: 06/22/2017
VIOLATION CODE: DUI
ARREST TIME: 00:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:42

SUBJECT'S LAST NAME: PAOLINI

DATE/TIME OF TEST: 09/22/2016 01:20:04

Diagnostics Check		OK	01:12
Air Blank		0.000	01:13
Control Test		0.076	01:13
Air Blank		0.000	01:13
Subject Sample	# 1	0.099	01:14
Breath Volume		1.628	01:14
Air Blank		0.000	01:15
Air Blank		0.000	01:17
Subject Sample	# 2	0.100	01:18
Breath Volume		1.335	01:18
Air Blank		0.000	01:19
Control Test		0.078	01:19
Air Blank		0.000	01:19
Diagnostics Check		OK	01:20

Data Download Date/Time: 10/24/2016 12:14

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESCOBAR,ALEJANDRO,
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf
ARRESTING OFFICER: ESCOBAR,ALEJANDRO
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf
LAST AGENCY INSPECTION: 09/20/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1698962
DRY GAS EXPIRY DATE: 06/22/2017
VIOLATION CODE: DUI
ARREST TIME: 19:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: DUDLEY

DATE/TIME OF TEST: 09/25/2016 20:35:02

Diagnostics Check		OK	20:32	
Air Blank		0.000	20:32	
Control Test		0.077	20:33	
Air Blank		0.000	20:33	
Subject Sample	# 1		20:33	Subject Test Refused
Breath Volume		0.000	20:33	
Air Blank		0.000	20:34	
Control Test		0.078	20:34	
Air Blank		0.000	20:34	
Diagnostics Check		OK	20:35	

Data Download Date/Time: 10/24/2016 12:14

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESCOBAR,ALEJANDRO,
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf
ARRESTING OFFICER: ESCOBAR,ALEJANDRO
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf
LAST AGENCY INSPECTION: 09/20/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1698962
DRY GAS EXPIRY DATE: 06/22/2017
VIOLATION CODE: DUI
ARREST TIME: 03:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:26

SUBJECT'S LAST NAME: GRADY

DATE/TIME OF TEST: 09/26/2016 04:06:10

Diagnostics Check		OK	04:03	
Air Blank		0.000	04:03	
Control Test		0.077	04:04	
Air Blank		0.000	04:04	
Subject Sample	# 1		04:04	Subject Test Refused
Breath Volume		0.000	04:04	
Air Blank		0.000	04:05	
Control Test		0.078	04:05	
Air Blank		0.000	04:06	
Diagnostics Check		OK	04:06	

Data Download Date/Time: 10/24/2016 12:14

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESCOBAR,ALEJANDRO,
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf
ARRESTING OFFICER: ESCOBAR,ALEJANDRO
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf
LAST AGENCY INSPECTION: 09/20/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1698962
DRY GAS EXPIRY DATE: 06/22/2017
VIOLATION CODE: DUI
ARREST TIME: 00:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:31

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 09/27/2016 01:07:45

Diagnostics Check		OK	01:01
Air Blank		0.000	01:01
Control Test		0.077	01:01
Air Blank		0.000	01:02
Subject Sample	# 1	0.205	01:02
Breath Volume		1.195	01:02
Air Blank		0.000	01:03
Air Blank		0.000	01:05
Subject Sample	# 2	0.208	01:06
Breath Volume		1.796	01:06
Air Blank		0.000	01:06
Control Test		0.077	01:07
Air Blank		0.000	01:07
Diagnostics Check		OK	01:07

Data Download Date/Time: 10/24/2016 12:14

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KU,CHRIS,U
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf
ARRESTING OFFICER: CAPALDI,JUSTIN
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf
LAST AGENCY INSPECTION: 09/20/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1698962
DRY GAS EXPIRY DATE: 06/22/2017
VIOLATION CODE: DUI
ARREST TIME: 21:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: CARDONA

DATE/TIME OF TEST: 09/28/2016 22:47:53

Diagnostics Check		OK	22:45	
Air Blank		0.000	22:45	
Control Test		0.077	22:45	
Air Blank		0.000	22:46	
Subject Sample	# 1		22:46	Subject Test Refused
Breath Volume		0.000	22:46	
Air Blank		0.000	22:46	
Control Test		0.078	22:47	
Air Blank		0.000	22:47	
Diagnostics Check		OK	22:47	

Data Download Date/Time: 10/24/2016 12:14

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAKER,TYLER,R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: GRAY,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/31/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 16:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:45

SUBJECT'S LAST NAME: ;6360102342052458201

DATE/TIME OF TEST: 09/01/2016 17:14:38

Diagnostics Check		OK	17:05
Air Blank		0.000	17:05
Control Test		0.079	17:06
Air Blank		0.000	17:06
Subject Sample	# 1	0.199	17:07
Breath Volume		2.027	17:07
Air Blank		0.000	17:08
Air Blank		0.000	17:09
Subject Sample	# 2	0.190	17:12
Breath Volume		1.332	17:12
Air Blank		0.000	17:13
Control Test		0.078	17:13
Air Blank		0.000	17:14
Diagnostics Check		OK	17:14

Data Download Date/Time: 09/14/2016 13:11

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBER,JOSEPH,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BROWN,T.A.
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/31/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:09

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 09/07/2016 01:36:04

Diagnostics Check		OK	01:29
Air Blank		0.000	01:30
Control Test		0.079	01:30
Air Blank		0.000	01:30
Subject Sample	# 1	0.158	01:31
Breath Volume		3.824	01:31
Air Blank		0.000	01:32
Air Blank		0.000	01:33
Subject Sample	# 2	0.156	01:34
Breath Volume		4.191	01:34
Air Blank		0.000	01:35
Control Test		0.079	01:35
Air Blank		0.000	01:35
Diagnostics Check		OK	01:36

Data Download Date/Time: 09/14/2016 13:11

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON,BRADEN,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHITE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/31/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: BLUMENAUER

DATE/TIME OF TEST: 09/09/2016 02:42:24

Diagnostics Check		OK	02:36
Air Blank		0.000	02:36
Control Test		0.078	02:36
Air Blank		0.000	02:37
Subject Sample	# 1	0.173	02:37
Breath Volume		2.171	02:37
Air Blank		0.000	02:38
Air Blank		0.000	02:40
Subject Sample	# 2	0.166	02:40
Breath Volume		1.171	02:40
Air Blank		0.000	02:41
Control Test		0.078	02:41
Air Blank		0.000	02:42
Diagnostics Check		OK	02:42

Data Download Date/Time: 09/14/2016 13:12

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON,BRADEN,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CAMACHO,K
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 08/31/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:26

SUBJECT'S LAST NAME: MISKANIC

DATE/TIME OF TEST: 09/09/2016 04:53:42

Diagnostics Check		OK	04:46
Air Blank		0.000	04:47
Control Test		0.079	04:47
Air Blank		0.000	04:48
Subject Sample	# 1	0.161	04:48
Breath Volume		1.421	04:48
Air Blank		0.000	04:49
Air Blank		0.000	04:51
Subject Sample	# 2	0.172	04:52
Breath Volume		1.175	04:52
Air Blank		0.000	04:52
Control Test		0.079	04:53
Air Blank		0.000	04:53
Diagnostics Check		OK	04:53

Data Download Date/Time: 09/14/2016 13:12

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRUETT,RICHARD,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: GARCIA,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/31/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: JORDAN

DATE/TIME OF TEST: 09/10/2016 03:29:37

Diagnostics Check		OK	03:26	
Air Blank		0.000	03:27	
Control Test		0.078	03:27	
Air Blank		0.000	03:28	
Subject Sample	# 1		03:28	Subject Test Refused
Breath Volume		0.000	03:28	
Air Blank		0.000	03:28	
Control Test		0.078	03:29	
Air Blank		0.000	03:29	
Diagnostics Check		OK	03:29	

Data Download Date/Time: 09/14/2016 13:12

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPICER,LORETHA,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SAMES,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/31/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 06:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:16

SUBJECT'S LAST NAME: BRANNAN

DATE/TIME OF TEST: 09/10/2016 08:43:35

Diagnostics Check		OK	08:36
Air Blank		0.000	08:37
Control Test		0.078	08:37
Air Blank		0.000	08:37
Subject Sample	# 1	0.223	08:38
Breath Volume		1.824	08:38
Air Blank		0.000	08:39
Air Blank		0.000	08:41
Subject Sample	# 2	0.212	08:41
Breath Volume		1.285	08:41
Air Blank		0.000	08:42
Control Test		0.077	08:42
Air Blank		0.000	08:43
Diagnostics Check		OK	08:43

Data Download Date/Time: 09/14/2016 13:12

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRUETT,RICHARD,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: GARCIA,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/31/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: ELLIS

DATE/TIME OF TEST: 09/12/2016 00:41:57

Diagnostics Check		OK	00:35
Air Blank		0.000	00:36
Control Test		0.078	00:36
Air Blank		0.000	00:37
Subject Sample	# 1	0.000	00:37
Breath Volume		1.300	00:37
Air Blank		0.000	00:37
Air Blank		0.000	00:39
Subject Sample	# 2	0.000	00:40
Breath Volume		2.035	00:40
Air Blank		0.000	00:41
Control Test		0.079	00:41
Air Blank		0.000	00:41
Diagnostics Check		OK	00:41

Data Download Date/Time: 09/14/2016 13:12

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTELL,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BROWN,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 09/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 16:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:15

SUBJECT'S LAST NAME: GUTIERREZ

DATE/TIME OF TEST: 09/14/2016 18:41:54

Diagnostics Check		OK	18:35
Air Blank		0.000	18:35
Control Test		0.078	18:36
Air Blank		0.000	18:36
Subject Sample	# 1	0.035	18:37
Breath Volume		1.351	18:37
Air Blank		0.000	18:37
Air Blank		0.000	18:39
Subject Sample	# 2	0.033	18:40
Breath Volume		2.765	18:40
Air Blank		0.000	18:40
Control Test		0.077	18:41
Air Blank		0.000	18:41
Diagnostics Check		OK	18:41

Data Download Date/Time: 09/29/2016 14:09

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTELL,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: THOMAS,C
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 09/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:01

SUBJECT'S LAST NAME: RAISON

DATE/TIME OF TEST: 09/14/2016 22:24:54

Diagnostics Check		OK	22:22	
Air Blank		0.000	22:22	
Control Test		0.078	22:22	
Air Blank		0.000	22:23	
Subject Sample	# 1		22:23	Subject Test Refused
Breath Volume		0.000	22:23	
Air Blank		0.000	22:23	
Control Test		0.078	22:24	
Air Blank		0.000	22:24	
Diagnostics Check		OK	22:24	

Data Download Date/Time: 09/29/2016 14:09

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTELL,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: KLING,D
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 09/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:21

SUBJECT'S LAST NAME: BLAKE

DATE/TIME OF TEST: 09/15/2016 21:50:26

Diagnostics Check		OK	21:41
Air Blank		0.000	21:41
Control Test		0.078	21:42
Air Blank		0.000	21:42
Subject Sample	# 1	0.158	21:45
Breath Volume		1.929	21:45
Air Blank		0.000	21:46
Air Blank		0.000	21:48
Subject Sample	# 2	0.163	21:48
Breath Volume		1.957	21:48
Air Blank		0.000	21:49
Control Test		0.076	21:49
Air Blank		0.000	21:50
Diagnostics Check		OK	21:50

Data Download Date/Time: 09/29/2016 14:09

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRUETT,RICHARD,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHITE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 09/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: BATES

DATE/TIME OF TEST: 09/19/2016 04:28:48

Diagnostics Check		OK	04:22
Air Blank		0.000	04:22
Control Test		0.079	04:23
Air Blank		0.000	04:23
Subject Sample	# 1	0.134	04:24
Breath Volume		1.238	04:24
Air Blank		0.000	04:24
Air Blank		0.000	04:26
Subject Sample	# 2	0.134	04:27
Breath Volume		1.542	04:27
Air Blank		0.000	04:27
Control Test		0.078	04:28
Air Blank		0.000	04:28
Diagnostics Check		OK	04:28

Data Download Date/Time: 09/29/2016 14:09

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRUETT,RICHARD,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: RODRIGUEZ,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 09/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:51

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 09/22/2016 01:24:26

Diagnostics Check		OK	01:17
Air Blank		0.000	01:17
Control Test		0.078	01:17
Air Blank		0.000	01:18
Subject Sample	# 1	0.134	01:19
Breath Volume		1.464	01:19
Air Blank		0.000	01:20
Air Blank		0.000	01:21
Subject Sample	# 2	0.129	01:22
Breath Volume		1.429	01:22
Air Blank		0.000	01:23
Control Test		0.077	01:23
Air Blank		0.000	01:24
Diagnostics Check		OK	01:24

Data Download Date/Time: 09/29/2016 14:09

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAUSEY,JOHN,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BROWN,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 09/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: DUMAS

DATE/TIME OF TEST: 09/24/2016 21:57:32

Diagnostics Check		OK	21:50
Air Blank		0.000	21:51
Control Test		0.078	21:51
Air Blank		0.000	21:52
Subject Sample	# 1	0.141	21:52
Breath Volume		1.878	21:52
Air Blank		0.000	21:53
Air Blank		0.000	21:55
Subject Sample	# 2	0.132	21:55
Breath Volume		2.148	21:55
Air Blank		0.000	21:56
Control Test		0.078	21:56
Air Blank		0.000	21:57
Diagnostics Check		OK	21:57

Data Download Date/Time: 09/29/2016 14:09

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHEPHERD,PAULINE,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MARRERO,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 09/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: HARTLEY

DATE/TIME OF TEST: 09/25/2016 22:18:53

Diagnostics Check		OK	22:10
Air Blank		0.000	22:11
Control Test		0.078	22:11
Air Blank		0.000	22:11
Subject Sample	# 1	0.120	22:14
Breath Volume		2.574	22:14
Air Blank		0.000	22:14
Air Blank		0.000	22:16
Subject Sample	# 2	0.119	22:17
Breath Volume		1.867	22:17
Air Blank		0.000	22:17
Control Test		0.077	22:18
Air Blank		0.000	22:18
Diagnostics Check		OK	22:18

Data Download Date/Time: 09/29/2016 14:09

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON,B,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MOSLEY,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 09/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: OAKS

DATE/TIME OF TEST: 09/26/2016 00:26:50

Diagnostics Check		OK	00:20
Air Blank		0.000	00:20
Control Test		0.077	00:21
Air Blank		0.000	00:21
Subject Sample	# 1	0.177	00:22
Breath Volume		3.203	00:22
Air Blank		0.000	00:22
Air Blank		0.000	00:24
Subject Sample	# 2	0.181	00:25
Breath Volume		2.812	00:25
Air Blank		0.000	00:25
Control Test		0.077	00:26
Air Blank		0.000	00:26
Diagnostics Check		OK	00:26

Data Download Date/Time: 09/29/2016 14:09

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON,BRADEN,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: KERSHAW,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 09/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:28

SUBJECT'S LAST NAME: PERRY

DATE/TIME OF TEST: 09/26/2016 02:55:28

Diagnostics Check		OK	02:48
Air Blank		0.000	02:49
Control Test		0.078	02:49
Air Blank		0.000	02:49
Subject Sample	# 1	0.131	02:50
Breath Volume		1.183	02:50
Air Blank		0.000	02:51
Air Blank		0.000	02:53
Subject Sample	# 2	0.139	02:53
Breath Volume		1.714	02:53
Air Blank		0.000	02:54
Control Test		0.078	02:54
Air Blank		0.000	02:55
Diagnostics Check		OK	02:55

Data Download Date/Time: 09/29/2016 14:09

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTELL,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: EANES,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 09/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:47

SUBJECT'S LAST NAME: MIRON

DATE/TIME OF TEST: 09/26/2016 22:14:32

Diagnostics Check		OK	22:07
Air Blank		0.000	22:08
Control Test		0.078	22:08
Air Blank		0.000	22:08
Subject Sample	# 1	0.269	22:09
Breath Volume		4.179	22:09
Air Blank		0.000	22:10
Air Blank		0.000	22:12
Subject Sample	# 2	0.260	22:12
Breath Volume		4.019	22:12
Air Blank		0.000	22:13
Control Test		0.077	22:13
Air Blank		0.000	22:14
Diagnostics Check		OK	22:14

Data Download Date/Time: 09/29/2016 14:09

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTELL,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HOWARD,J
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 09/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:22

SUBJECT'S LAST NAME: REINOLDS

DATE/TIME OF TEST: 09/27/2016 21:48:56

Diagnostics Check		OK	21:42
Air Blank		0.000	21:43
Control Test		0.079	21:43
Air Blank		0.000	21:43
Subject Sample	# 1	0.000	21:44
Breath Volume		1.433	21:44
Air Blank		0.000	21:44
Air Blank		0.000	21:46
Subject Sample	# 2	0.000	21:47
Breath Volume		2.226	21:47
Air Blank		0.000	21:48
Control Test		0.079	21:48
Air Blank		0.000	21:48
Diagnostics Check		OK	21:48

Data Download Date/Time: 09/29/2016 14:09

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: STANLEY,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 09/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: CAMP

DATE/TIME OF TEST: 09/28/2016 01:55:30

Diagnostics Check		OK	01:52	
Air Blank		0.000	01:53	
Control Test		0.079	01:53	
Air Blank		0.000	01:53	
Subject Sample	# 1		01:54	Subject Test Refused
Breath Volume		0.000	01:54	
Air Blank		0.000	01:54	
Control Test		0.078	01:54	
Air Blank		0.000	01:55	
Diagnostics Check		OK	01:55	

Data Download Date/Time: 09/29/2016 14:09

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRUETT,RICHARD,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: THOMPSON,M
ARRESTING OFFICER AGENCY: Atlantic Beach PD
LAST AGENCY INSPECTION: 09/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:02

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 09/28/2016 04:29:28

Diagnostics Check		OK	04:22
Air Blank		0.000	04:23
Control Test		0.079	04:23
Air Blank		0.000	04:24
Subject Sample	# 1	0.180	04:24
Breath Volume		2.195	04:24
Air Blank		0.000	04:25
Air Blank		0.000	04:27
Subject Sample	# 2	0.177	04:27
Breath Volume		1.777	04:27
Air Blank		0.000	04:28
Control Test		0.078	04:28
Air Blank		0.000	04:29
Diagnostics Check		OK	04:29

Data Download Date/Time: 09/29/2016 14:09

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBER,JOSEPH,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HEALY,G
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 09/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: HOOVER

DATE/TIME OF TEST: 09/29/2016 03:22:45

Diagnostics Check		OK	03:15
Air Blank		0.000	03:16
Control Test		0.078	03:16
Air Blank		0.000	03:17
Subject Sample	# 1	0.140	03:17
Breath Volume		1.402	03:17
Air Blank		0.000	03:18
Air Blank		0.000	03:20
Subject Sample	# 2	0.151	03:21
Breath Volume		1.585	03:21
Air Blank		0.000	03:21
Control Test		0.078	03:22
Air Blank		0.000	03:22
Diagnostics Check		OK	03:22

Data Download Date/Time: 09/29/2016 14:09

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINSON,C,A
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: FORNASH,A
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 08/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: LEE

DATE/TIME OF TEST: 09/03/2016 02:46:37

Diagnostics Check		OK	02:35	
Air Blank		0.000	02:36	
Control Test		0.082	02:36	
Air Blank		0.000	02:36	
Subject Sample	# 1		02:40	No Sample Provided
Breath Volume		0.000	02:40	
Air Blank		0.000	02:40	
Air Blank		0.000	02:42	
Subject Sample	# 2		02:45	Subject Test Refused
Breath Volume		0.000	02:45	
Air Blank		0.000	02:45	
Control Test		0.082	02:46	
Air Blank		0.000	02:46	
Diagnostics Check		OK	02:46	

Data Download Date/Time: 09/25/2016 18:51

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ASHE,KEVIN,B
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: ENGLISH,S
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 08/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: CAMERON

DATE/TIME OF TEST: 09/06/2016 01:31:45

Diagnostics Check		OK	01:25
Air Blank		0.000	01:26
Control Test		0.081	01:26
Air Blank		0.000	01:26
Subject Sample	# 1	0.249	01:27
Breath Volume		1.277	01:27
Air Blank		0.000	01:27
Air Blank		0.000	01:29
Subject Sample	# 2	0.237	01:30
Breath Volume		2.156	01:30
Air Blank		0.000	01:30
Control Test		0.081	01:31
Air Blank		0.000	01:31
Diagnostics Check		OK	01:31

Data Download Date/Time: 09/25/2016 18:51

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSON,MATTHEW,P
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: LLEWELLYN,D
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 08/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:18

SUBJECT'S LAST NAME: OWEN

DATE/TIME OF TEST: 09/11/2016 04:50:20

Diagnostics Check		OK	04:42
Air Blank		0.000	04:43
Control Test		0.081	04:43
Air Blank		0.000	04:44
Subject Sample	# 1	0.228	04:45
Breath Volume		3.890	04:45
Air Blank		0.000	04:46
Air Blank		0.000	04:48
Subject Sample	# 2	0.214	04:48
Breath Volume		2.832	04:48
Air Blank		0.000	04:49
Control Test		0.080	04:49
Air Blank		0.000	04:50
Diagnostics Check		OK	04:50

Data Download Date/Time: 09/25/2016 18:51

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINSON,CODY,A
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: SAUNDERS,J
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 08/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 17:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:20

SUBJECT'S LAST NAME: NADAL

DATE/TIME OF TEST: 09/12/2016 19:03:30

Diagnostics Check		OK	18:46	
Air Blank		0.000	18:47	
Control Test		0.081	18:47	
Air Blank		0.000	18:48	
Subject Sample	# 1	0.397	18:50	
Breath Volume		1.269	18:50	
Air Blank		0.000	18:51	
Air Blank		0.000	18:53	
Subject Sample	# 2		18:56	No Sample Provided
Breath Volume		0.000	18:56	
Air Blank		0.000	18:56	
Air Blank		0.000	18:58	
Subject Sample	# 3		19:01	No Sample Provided
Breath Volume		0.000	19:01	
Air Blank		0.000	19:02	
Control Test		0.080	19:02	
Air Blank		0.000	19:03	
Diagnostics Check		OK	19:03	

Data Download Date/Time: 09/25/2016 18:51

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ASHE,KEVIN,B
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: WHITE,M
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 09/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:23

SUBJECT'S LAST NAME: GLOWACZ

DATE/TIME OF TEST: 09/30/2016 02:49:43

Diagnostics Check		OK	02:43
Air Blank		0.000	02:43
Control Test		0.082	02:44
Air Blank		0.000	02:44
Subject Sample	# 1	0.154	02:45
Breath Volume		3.289	02:45
Air Blank		0.000	02:45
Air Blank		0.000	02:47
Subject Sample	# 2	0.144	02:48
Breath Volume		2.945	02:48
Air Blank		0.000	02:48
Control Test		0.082	02:49
Air Blank		0.000	02:49
Diagnostics Check		OK	02:49

Data Download Date/Time: 10/09/2016 09:46

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: WILLIAMS,SEAN
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 09/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 22:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:43

SUBJECT'S LAST NAME: IRACE

DATE/TIME OF TEST: 09/24/2016 23:10:30

Diagnostics Check		OK	23:04
Air Blank		0.000	23:04
Control Test		0.080	23:05
Air Blank		0.000	23:05
Subject Sample	# 1	0.157	23:06
Breath Volume		3.535	23:06
Air Blank		0.000	23:06
Air Blank		0.000	23:08
Subject Sample	# 2	0.151	23:09
Breath Volume		2.707	23:09
Air Blank		0.000	23:09
Control Test		0.079	23:09
Air Blank		0.000	23:10
Diagnostics Check		OK	23:10

Data Download Date/Time: 10/19/2016 11:47

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: DIAZ,DANNY
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 09/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 02:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:29

SUBJECT'S LAST NAME: RAMOS COMPANIONI

DATE/TIME OF TEST: 09/25/2016 02:57:23

Diagnostics Check		OK	02:51
Air Blank		0.000	02:51
Control Test		0.080	02:52
Air Blank		0.000	02:52
Subject Sample	# 1	0.123	02:52
Breath Volume		3.964	02:52
Air Blank		0.000	02:53
Air Blank		0.000	02:55
Subject Sample	# 2	0.126	02:55
Breath Volume		3.867	02:55
Air Blank		0.000	02:56
Control Test		0.080	02:56
Air Blank		0.000	02:57
Diagnostics Check		OK	02:57

Data Download Date/Time: 10/19/2016 11:47

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SERPHOS,STEVEN
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 09/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 03:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:18

SUBJECT'S LAST NAME: JOHNSON JR

DATE/TIME OF TEST: 09/25/2016 03:47:04

Diagnostics Check		OK	03:40
Air Blank		0.000	03:41
Control Test		0.081	03:41
Air Blank		0.000	03:42
Subject Sample	# 1	0.309	03:42
Breath Volume		2.285	03:42
Air Blank		0.000	03:43
Air Blank		0.000	03:45
Subject Sample	# 2	0.301	03:45
Breath Volume		2.757	03:45
Air Blank		0.000	03:46
Control Test		0.080	03:46
Air Blank		0.000	03:46
Diagnostics Check		OK	03:47

Data Download Date/Time: 10/19/2016 11:47

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SAPP,JOSHUA
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 09/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 21:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:22

SUBJECT'S LAST NAME: ESCOBAR

DATE/TIME OF TEST: 09/25/2016 22:51:15

Diagnostics Check		OK	22:45
Air Blank		0.000	22:45
Control Test		0.080	22:45
Air Blank		0.000	22:46
Subject Sample	# 1	0.037	22:46
Breath Volume		1.925	22:46
Air Blank		0.000	22:47
Air Blank		0.000	22:49
Subject Sample	# 2	0.036	22:49
Breath Volume		2.605	22:49
Air Blank		0.000	22:50
Control Test		0.080	22:50
Air Blank		0.000	22:51
Diagnostics Check		OK	22:51

Data Download Date/Time: 10/19/2016 11:47

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SERPHOS,STEVEN
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 09/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 02:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:46

SUBJECT'S LAST NAME: TRAN

DATE/TIME OF TEST: 09/26/2016 03:14:40

Diagnostics Check		OK	03:08
Air Blank		0.000	03:09
Control Test		0.080	03:09
Air Blank		0.000	03:09
Subject Sample	# 1	0.070	03:10
Breath Volume		1.960	03:10
Air Blank		0.000	03:10
Air Blank		0.000	03:12
Subject Sample	# 2	0.058	03:13
Breath Volume		1.125	03:13
Air Blank		0.000	03:13
Control Test		0.080	03:14
Air Blank		0.000	03:14
Diagnostics Check		OK	03:14

Data Download Date/Time: 10/19/2016 11:47

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SAPP,JOSHUA
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 09/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 19:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:15

SUBJECT'S LAST NAME: HATEM

DATE/TIME OF TEST: 09/30/2016 20:55:54

Diagnostics Check		OK	20:49
Air Blank		0.000	20:49
Control Test		0.080	20:50
Air Blank		0.000	20:50
Subject Sample	# 1	0.000	20:51
Breath Volume		2.285	20:51
Air Blank		0.000	20:51
Air Blank		0.000	20:53
Subject Sample	# 2	0.000	20:54
Breath Volume		1.757	20:54
Air Blank		0.000	20:55
Control Test		0.080	20:55
Air Blank		0.000	20:55
Diagnostics Check		OK	20:55

Data Download Date/Time: 10/19/2016 11:47

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALVARADO,ANGEL,L
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: ALVARADO,ANGEL
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 08/19/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A1
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 19:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:31

SUBJECT'S LAST NAME: STEPHENS

DATE/TIME OF TEST: 09/01/2016 20:58:06

Diagnostics Check		OK	20:53	
Air Blank		0.000	20:53	
Control Test		0.079	20:53	
Air Blank		0.000	20:54	
Subject Sample	# 1		20:56	Subject Test Refused
Breath Volume		0.000	20:56	
Air Blank		0.000	20:57	
Control Test		0.081	20:57	
Air Blank		0.000	20:58	
Diagnostics Check		OK	20:58	

Data Download Date/Time: 09/15/2016 10:19

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALVARADO,ANGEL,L
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: ALVARADO,ANGEL
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 08/19/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A1
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: WRIGHT

DATE/TIME OF TEST: 09/09/2016 22:03:51

Diagnostics Check		OK	21:52	
Air Blank		0.000	21:52	
Control Test		0.080	21:53	
Air Blank		0.000	21:53	
Subject Sample	# 1		21:56	No Sample Provided
Breath Volume		0.000	21:56	
Air Blank		0.000	21:57	
Air Blank		0.000	21:59	
Subject Sample	# 2		22:02	No Sample Provided
Breath Volume		0.000	22:02	
Air Blank		0.000	22:02	
Control Test		0.080	22:03	
Air Blank		0.000	22:03	
Diagnostics Check		OK	22:03	

Data Download Date/Time: 09/15/2016 10:19

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALVARADO,ANGEL,L
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: ALVARADO,ANGEL
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 08/19/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A1
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: WRIGHT

DATE/TIME OF TEST: 09/09/2016 22:18:37

Diagnostics Check		OK	22:10
Air Blank		0.000	22:10
Control Test		0.080	22:10
Air Blank		0.000	22:11
Subject Sample	# 1	0.159	22:13
Breath Volume		1.781	22:13
Air Blank		0.000	22:14
Air Blank		0.000	22:16
Subject Sample	# 2	0.152	22:17
Breath Volume		1.457	22:17
Air Blank		0.000	22:17
Control Test		0.079	22:18
Air Blank		0.000	22:18
Diagnostics Check		OK	22:18

Data Download Date/Time: 09/15/2016 10:19

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,WILLIAM,T
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SMITH,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 08/19/2016

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 03415080A1
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:38

SUBJECT'S LAST NAME: BROOME

DATE/TIME OF TEST: 09/10/2016 04:08:35

Diagnostics Check		OK	04:00
Air Blank		0.000	04:00
Control Test		0.079	04:01
Air Blank		0.000	04:01
Subject Sample	# 1	0.173	04:04
Breath Volume		1.242	04:04
Air Blank		0.000	04:04
Air Blank		0.000	04:06
Subject Sample	# 2	0.179	04:07
Breath Volume		1.246	04:07
Air Blank		0.000	04:07
Control Test		0.078	04:08
Air Blank		0.000	04:08
Diagnostics Check		OK	04:08

Data Download Date/Time: 09/15/2016 10:19

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRIOLAT,CASEY,J
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: BRIOLAT,CASEY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 08/19/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A1
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 18:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:45

SUBJECT'S LAST NAME: CHEKRYGIN

DATE/TIME OF TEST: 09/10/2016 20:14:10

Diagnostics Check		OK	20:06
Air Blank		0.000	20:07
Control Test		0.080	20:07
Air Blank		0.000	20:07
Subject Sample	# 1	0.000	20:09
Breath Volume		2.324	20:09
Air Blank		0.000	20:09
Air Blank		0.000	20:11
Subject Sample	# 2	0.000	20:12
Breath Volume		2.078	20:12
Air Blank		0.000	20:13
Control Test		0.080	20:13
Air Blank		0.000	20:14
Diagnostics Check		OK	20:14

Data Download Date/Time: 09/15/2016 10:19

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARROW,MICHAEL,A
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: CROFT,CHARLES
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 08/19/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A1
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 18:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:30

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 09/11/2016 19:37:28

Diagnostics Check		OK	19:34	
Air Blank		0.000	19:34	
Control Test		0.080	19:34	
Air Blank		0.000	19:35	
Subject Sample	# 1		19:35	Subject Test Refused
Breath Volume		0.000	19:35	
Air Blank		0.000	19:36	
Control Test		0.079	19:36	
Air Blank		0.000	19:37	
Diagnostics Check		OK	19:37	

Data Download Date/Time: 09/15/2016 10:19

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALVARADO,ANGEL,L
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: WILLIAMS,C
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 08/19/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A1
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:13

SUBJECT'S LAST NAME: SCAVIOLA

DATE/TIME OF TEST: 09/11/2016 22:40:45

Diagnostics Check		OK	22:34
Air Blank		0.000	22:34
Control Test		0.080	22:34
Air Blank		0.000	22:35
Subject Sample	# 1	0.155	22:36
Breath Volume		2.386	22:36
Air Blank		0.000	22:36
Air Blank		0.000	22:38
Subject Sample	# 2	0.151	22:39
Breath Volume		1.683	22:39
Air Blank		0.000	22:39
Control Test		0.079	22:40
Air Blank		0.000	22:40
Diagnostics Check		OK	22:40

Data Download Date/Time: 09/15/2016 10:19

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALVARADO,ANGEL,L
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: ALVARADO,ANGEL
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 08/19/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A1
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: CHIMELIS

DATE/TIME OF TEST: 09/12/2016 02:06:16

Diagnostics Check		OK	02:01	
Air Blank		0.000	02:02	
Control Test		0.079	02:02	
Air Blank		0.000	02:03	
Subject Sample	# 1		02:04	Subject Test Refused
Breath Volume		0.000	02:04	
Air Blank		0.000	02:05	
Control Test		0.080	02:05	
Air Blank		0.000	02:06	
Diagnostics Check		OK	02:06	

Data Download Date/Time: 09/15/2016 10:19

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAUS,SCOTT,G
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: MAUS,SCOTT
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 08/19/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A1
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 15:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:04

SUBJECT'S LAST NAME: PEARSON

DATE/TIME OF TEST: 09/13/2016 17:30:52

Diagnostics Check		OK	17:24
Air Blank		0.000	17:24
Control Test		0.080	17:25
Air Blank		0.000	17:25
Subject Sample	# 1	0.259	17:26
Breath Volume		1.234	17:26
Air Blank		0.000	17:27
Air Blank		0.000	17:28
Subject Sample	# 2	0.253	17:29
Breath Volume		1.828	17:29
Air Blank		0.000	17:29
Control Test		0.080	17:30
Air Blank		0.000	17:30
Diagnostics Check		OK	17:30

Data Download Date/Time: 09/15/2016 10:19

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALVARADO,ANGEL,L
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: ALVARADO,ANGEL
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 09/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A1
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: JOHNES

DATE/TIME OF TEST: 09/16/2016 04:03:12

Diagnostics Check		OK	04:01	
Air Blank		0.000	04:01	
Control Test		0.078	04:01	
Air Blank		0.000	04:02	
Subject Sample	# 1		04:02	Sequence Aborted
Breath Volume		0.000	04:02	
Air Blank		0.000	04:03	

Data Download Date/Time: 10/14/2016 08:33

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MIDDLETON,DANIEL,S
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: MIDDLETON,DANIEL
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 09/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A1
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:13

SUBJECT'S LAST NAME: MALONE

DATE/TIME OF TEST: 09/24/2016 23:41:19

Diagnostics Check		OK	23:34
Air Blank		0.000	23:35
Control Test		0.078	23:35
Air Blank		0.000	23:36
Subject Sample	# 1	0.126	23:36
Breath Volume		2.664	23:36
Air Blank		0.000	23:37
Air Blank		0.000	23:38
Subject Sample	# 2	0.125	23:39
Breath Volume		2.515	23:39
Air Blank		0.000	23:40
Control Test		0.080	23:40
Air Blank		0.000	23:41
Diagnostics Check		OK	23:41

Data Download Date/Time: 10/14/2016 08:33

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001050

INSTRUMENT REGISTERED TO: CORAL SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HILDEBRANDT,N,E
BREATH TEST OPERATOR AGENCY: Coral Springs PD
ARRESTING OFFICER: MURDOCK,T
ARRESTING OFFICER AGENCY: Coral Springs PD
LAST AGENCY INSPECTION: 08/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:42

SUBJECT'S LAST NAME: FAVILLA

DATE/TIME OF TEST: 09/03/2016 04:19:44

Diagnostics Check		OK	04:11
Air Blank		0.000	04:12
Control Test		0.079	04:12
Air Blank		0.000	04:13
Subject Sample	# 1	0.048	04:13
Breath Volume		2.785	04:13
Air Blank		0.000	04:14
Air Blank		0.000	04:15
Subject Sample	# 2	0.045	04:18
Breath Volume		1.148	04:18
Air Blank		0.000	04:18
Control Test		0.079	04:19
Air Blank		0.000	04:19
Diagnostics Check		OK	04:19

Data Download Date/Time: 09/28/2016 13:49

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001050

INSTRUMENT REGISTERED TO: CORAL SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WETHERINGTON,STEPHEN,L
BREATH TEST OPERATOR AGENCY: Coral Springs PD
ARRESTING OFFICER: WETHERINGTON,STEPHEN
ARRESTING OFFICER AGENCY: Coral Springs PD
LAST AGENCY INSPECTION: 08/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:52

SUBJECT'S LAST NAME: KOTOV

DATE/TIME OF TEST: 09/06/2016 01:23:07

Diagnostics Check		OK	01:16
Air Blank		0.000	01:16
Control Test		0.080	01:16
Air Blank		0.000	01:17
Subject Sample	# 1	0.157	01:17
Breath Volume		2.281	01:17
Air Blank		0.000	01:18
Air Blank		0.000	01:20
Subject Sample	# 2	0.156	01:21
Breath Volume		2.843	01:21
Air Blank		0.000	01:22
Control Test		0.080	01:22
Air Blank		0.000	01:23
Diagnostics Check		OK	01:23

Data Download Date/Time: 09/28/2016 13:49

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001050

INSTRUMENT REGISTERED TO: CORAL SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WETHERINGTON,STEPHEN,L
BREATH TEST OPERATOR AGENCY: Coral Springs PD
ARRESTING OFFICER: WETHERINGTON,STEPHEN
ARRESTING OFFICER AGENCY: Coral Springs PD
LAST AGENCY INSPECTION: 08/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:46

SUBJECT'S LAST NAME: PIERRE

DATE/TIME OF TEST: 09/26/2016 04:14:10

Diagnostics Check	OK	04:12	
Air Blank	0.000	04:13	
Control Test	0.079	04:13	
Air Blank		04:13	RFI Detect
Air Blank	0.000	04:14	

Data Download Date/Time: 09/28/2016 13:49

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001050

INSTRUMENT REGISTERED TO: CORAL SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WETHERINGTON,STEPHEN,L
BREATH TEST OPERATOR AGENCY: Coral Springs PD
ARRESTING OFFICER: WETHERINGTON,STEPHEN
ARRESTING OFFICER AGENCY: Coral Springs PD
LAST AGENCY INSPECTION: 08/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:46

SUBJECT'S LAST NAME: PIERRE

DATE/TIME OF TEST: 09/26/2016 04:22:03

Diagnostics Check		OK	04:16
Air Blank		0.000	04:16
Control Test		0.079	04:16
Air Blank		0.000	04:17
Subject Sample	# 1	0.000	04:17
Breath Volume		2.101	04:17
Air Blank		0.000	04:18
Air Blank		0.000	04:20
Subject Sample	# 2	0.000	04:20
Breath Volume		2.187	04:20
Air Blank		0.000	04:21
Control Test		0.079	04:21
Air Blank		0.000	04:21
Diagnostics Check		OK	04:22

Data Download Date/Time: 09/28/2016 13:49

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001050

INSTRUMENT REGISTERED TO: CORAL SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WETHERINGTON,STEPHEN,L
BREATH TEST OPERATOR AGENCY: Coral Springs PD
ARRESTING OFFICER: WETHERINGTON,STEPHEN
ARRESTING OFFICER AGENCY: Coral Springs PD
LAST AGENCY INSPECTION: 09/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:22

SUBJECT'S LAST NAME: HADDOCK SANCHEZ

DATE/TIME OF TEST: 09/28/2016 23:54:03

Diagnostics Check		OK	23:47
Air Blank		0.000	23:48
Control Test		0.079	23:48
Air Blank		0.000	23:48
Subject Sample	# 1	0.163	23:49
Breath Volume		3.496	23:49
Air Blank		0.000	23:49
Air Blank		0.000	23:51
Subject Sample	# 2	0.159	23:52
Breath Volume		2.734	23:52
Air Blank		0.000	23:53
Control Test		0.079	23:53
Air Blank		0.000	23:53
Diagnostics Check		OK	23:54

Data Download Date/Time: 10/26/2016 14:46

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001050

INSTRUMENT REGISTERED TO: CORAL SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BYRNE,GRETCHEN,
BREATH TEST OPERATOR AGENCY: Coral Springs PD
ARRESTING OFFICER: CABRERA,ALEX
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf
LAST AGENCY INSPECTION: 09/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 14:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:16

SUBJECT'S LAST NAME: WALKOWSKI

DATE/TIME OF TEST: 09/30/2016 16:19:41

Diagnostics Check		OK	16:12
Air Blank		0.000	16:13
Control Test		0.079	16:13
Air Blank		0.000	16:13
Subject Sample	# 1	0.168	16:15
Breath Volume		2.324	16:15
Air Blank		0.000	16:15
Air Blank		0.000	16:17
Subject Sample	# 2	0.167	16:18
Breath Volume		2.714	16:18
Air Blank		0.000	16:18
Control Test		0.078	16:19
Air Blank		0.000	16:19
Diagnostics Check		OK	16:19

Data Download Date/Time: 10/26/2016 14:46

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPENCER,SHANE,
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: BROWN,JOEL
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 08/19/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 06:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:52

SUBJECT'S LAST NAME: OVANDO

DATE/TIME OF TEST: 09/03/2016 07:19:43

Diagnostics Check		OK	07:13
Air Blank		0.000	07:13
Control Test		0.078	07:13
Air Blank		0.000	07:14
Subject Sample	# 1	0.265	07:15
Breath Volume		1.898	07:15
Air Blank		0.000	07:15
Air Blank		0.000	07:17
Subject Sample	# 2	0.249	07:18
Breath Volume		1.808	07:18
Air Blank		0.000	07:18
Control Test		0.080	07:19
Air Blank		0.000	07:19
Diagnostics Check		OK	07:19

Data Download Date/Time: 09/25/2016 07:36

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRACKETT,MARK,O
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: BRACKETT,MARK
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 08/19/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: ZERO TOL
ARREST TIME: 13:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:32

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 09/07/2016 14:05:46

Diagnostics Check		OK	13:59
Air Blank		0.000	13:59
Control Test		0.081	14:00
Air Blank		0.000	14:00
Subject Sample	# 1	0.000	14:01
Breath Volume		1.847	14:01
Air Blank		0.000	14:01
Air Blank		0.000	14:03
Subject Sample	# 2	0.000	14:04
Breath Volume		1.480	14:04
Air Blank		0.000	14:04
Control Test		0.080	14:05
Air Blank		0.000	14:05
Diagnostics Check		OK	14:05

Data Download Date/Time: 09/25/2016 07:36

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WORTH,K,J
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: WORTH,K
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 08/19/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: COURT ORD
ARREST TIME: 19:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:30

SUBJECT'S LAST NAME: NOLAND

DATE/TIME OF TEST: 09/07/2016 20:01:03

Diagnostics Check		OK	19:55
Air Blank		0.000	19:55
Control Test		0.081	19:55
Air Blank		0.000	19:56
Subject Sample	# 1	0.000	19:56
Breath Volume		2.726	19:56
Air Blank		0.000	19:57
Air Blank		0.000	19:59
Subject Sample	# 2	0.000	19:59
Breath Volume		3.402	19:59
Air Blank		0.000	20:00
Control Test		0.080	20:00
Air Blank		0.000	20:00
Diagnostics Check		OK	20:01

Data Download Date/Time: 09/25/2016 07:36

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WORTH,K,J
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: WORTH,K
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 08/19/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: COURT ORD
ARREST TIME: 20:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 09/07/2016 21:19:55

Diagnostics Check		OK	21:14
Air Blank		0.000	21:14
Control Test		0.079	21:14
Air Blank		0.000	21:15
Subject Sample	# 1	0.000	21:15
Breath Volume		1.535	21:15
Air Blank		0.000	21:16
Air Blank		0.000	21:18
Subject Sample	# 2	0.000	21:18
Breath Volume		1.867	21:18
Air Blank		0.000	21:19
Control Test		0.079	21:19
Air Blank		0.000	21:19
Diagnostics Check		OK	21:19

Data Download Date/Time: 09/25/2016 07:36

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRACKETT,MARK,O
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: BRACKETT,MARK
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 08/19/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: ZERO TOL
ARREST TIME: 11:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:39

SUBJECT'S LAST NAME: ROBINETTE

DATE/TIME OF TEST: 09/08/2016 12:12:39

Diagnostics Check		OK	12:06
Air Blank		0.000	12:06
Control Test		0.080	12:07
Air Blank		0.000	12:07
Subject Sample	# 1	0.000	12:08
Breath Volume		3.691	12:08
Air Blank		0.000	12:08
Air Blank		0.000	12:10
Subject Sample	# 2	0.000	12:11
Breath Volume		2.382	12:11
Air Blank		0.000	12:11
Control Test		0.079	12:12
Air Blank		0.000	12:12
Diagnostics Check		OK	12:12

Data Download Date/Time: 09/25/2016 07:36

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORSE,MICHAEL,P
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: RITENOUR,S
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 08/19/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:19

SUBJECT'S LAST NAME: MCCUMBER

DATE/TIME OF TEST: 09/10/2016 03:47:14

Diagnostics Check		OK	03:39
Air Blank		0.000	03:40
Control Test		0.078	03:40
Air Blank		0.000	03:41
Subject Sample	# 1	0.154	03:42
Breath Volume		3.835	03:42
Air Blank		0.000	03:43
Air Blank		0.000	03:45
Subject Sample	# 2	0.137	03:45
Breath Volume		2.351	03:45
Air Blank		0.000	03:46
Control Test		0.078	03:46
Air Blank		0.000	03:47
Diagnostics Check		OK	03:47

Data Download Date/Time: 09/25/2016 07:36

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORSE,MICHAEL,P
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: RITENOUR,S
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 08/19/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:16

SUBJECT'S LAST NAME: KELLEY

DATE/TIME OF TEST: 09/11/2016 03:43:12

Diagnostics Check		OK	03:36
Air Blank		0.000	03:36
Control Test		0.079	03:37
Air Blank		0.000	03:37
Subject Sample	# 1	0.209	03:38
Breath Volume		2.355	03:38
Air Blank		0.000	03:39
Air Blank		0.000	03:40
Subject Sample	# 2	0.213	03:41
Breath Volume		2.535	03:41
Air Blank		0.000	03:42
Control Test		0.080	03:42
Air Blank		0.000	03:43
Diagnostics Check		OK	03:43

Data Download Date/Time: 09/25/2016 07:36

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNELL,COREY,J
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: SNELL,COREY
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 08/19/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: COURT ORD
ARREST TIME: 17:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:45

SUBJECT'S LAST NAME: GUNN

DATE/TIME OF TEST: 09/20/2016 18:19:57

Diagnostics Check		OK	18:13
Air Blank		0.000	18:14
Control Test		0.079	18:14
Air Blank		0.000	18:15
Subject Sample	# 1	0.000	18:15
Breath Volume		1.824	18:15
Air Blank		0.000	18:16
Air Blank		0.000	18:18
Subject Sample	# 2	0.000	18:18
Breath Volume		1.746	18:18
Air Blank		0.000	18:19
Control Test		0.078	18:19
Air Blank		0.000	18:19
Diagnostics Check		OK	18:19

Data Download Date/Time: 09/25/2016 07:36

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUNN,TRAVIS,M
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: DUNN,TRAVIS
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 09/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: COURT ORD
ARREST TIME: 09:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:30

SUBJECT'S LAST NAME: CRAVEN

DATE/TIME OF TEST: 09/26/2016 10:16:41

Diagnostics Check		OK	10:10
Air Blank		0.000	10:11
Control Test		0.080	10:11
Air Blank		0.000	10:12
Subject Sample	# 1	0.000	10:12
Breath Volume		2.796	10:12
Air Blank		0.000	10:12
Air Blank		0.000	10:14
Subject Sample	# 2	0.000	10:15
Breath Volume		2.875	10:15
Air Blank		0.000	10:15
Control Test		0.079	10:16
Air Blank		0.000	10:16
Diagnostics Check		OK	10:16

Data Download Date/Time: 10/23/2016 09:59

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MELTZER,CHRISTOPHER,J
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: MELTZER,CHRISTOPHER
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 08/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:47

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 09/03/2016 04:22:42

Diagnostics Check		OK	04:19	
Air Blank		0.000	04:20	
Control Test		0.076	04:20	
Air Blank		0.000	04:20	
Subject Sample	# 1		04:21	Subject Test Refused
Breath Volume		0.000	04:21	
Air Blank		0.000	04:21	
Control Test		0.077	04:22	
Air Blank		0.000	04:22	
Diagnostics Check		OK	04:22	

Data Download Date/Time: 09/06/2016 12:06

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MELTZER,CHRISTOPHER,J
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: CHRISTELLO,CLIFFORD
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 08/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:15

SUBJECT'S LAST NAME: ROBERTS

DATE/TIME OF TEST: 09/03/2016 04:44:20

Diagnostics Check		OK	04:37
Air Blank		0.000	04:38
Control Test		0.077	04:38
Air Blank		0.000	04:39
Subject Sample	# 1	0.158	04:39
Breath Volume		1.871	04:39
Air Blank		0.000	04:40
Air Blank		0.000	04:42
Subject Sample	# 2	0.178	04:42
Breath Volume		2.750	04:42
Air Blank		0.000	04:43
Control Test		0.076	04:43
Air Blank		0.000	04:44
Diagnostics Check		OK	04:44

Data Download Date/Time: 09/06/2016 12:06

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MELTZER,CHRISTOPHER,J
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: MELTZER,CHRISTOPHER
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 08/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: COX

DATE/TIME OF TEST: 09/04/2016 03:09:05

Diagnostics Check		OK	03:01
Air Blank		0.000	03:02
Control Test		0.076	03:02
Air Blank		0.000	03:02
Subject Sample	# 1	0.138	03:04
Breath Volume		1.871	03:04
Air Blank		0.000	03:05
Air Blank		0.000	03:06
Subject Sample	# 2	0.138	03:07
Breath Volume		1.199	03:07
Air Blank		0.000	03:08
Control Test		0.076	03:08
Air Blank		0.000	03:09
Diagnostics Check		OK	03:09

Data Download Date/Time: 09/06/2016 12:06

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MELTZER,CHRISTOPHER,J
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: CHRISTELLO,CLIFFORD
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 08/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:31

SUBJECT'S LAST NAME: MORTAROTTI

DATE/TIME OF TEST: 09/04/2016 03:22:48

Diagnostics Check		OK	03:14
Air Blank		0.000	03:14
Control Test		0.076	03:15
Air Blank		0.000	03:15
Subject Sample	# 1	0.189	03:18
Breath Volume		1.730	03:18
Air Blank		0.000	03:18
Air Blank		0.000	03:20
Subject Sample	# 2	0.180	03:21
Breath Volume		1.550	03:21
Air Blank		0.000	03:21
Control Test		0.075	03:22
Air Blank		0.000	03:22
Diagnostics Check		OK	03:22

Data Download Date/Time: 09/06/2016 12:06

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUCAS,NATHANIEL,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: LUCAS,NATHANIEL
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 08/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: ROMUALDO

DATE/TIME OF TEST: 09/04/2016 03:58:00

Diagnostics Check		OK	03:47
Air Blank		0.000	03:48
Control Test		0.076	03:48
Air Blank		0.000	03:49
Subject Sample	# 1	0.128	03:52
Breath Volume		1.355	03:52
Air Blank		0.000	03:52
Air Blank		0.000	03:54
Subject Sample	# 2	0.134	03:56
Breath Volume		1.730	03:56
Air Blank		0.000	03:57
Control Test		0.075	03:57
Air Blank		0.000	03:57
Diagnostics Check		OK	03:58

Data Download Date/Time: 09/06/2016 12:06

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MELTZER,CHRISTOPHER,J
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: MELTZER,CHRISTOPHER
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 08/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 18:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:57

SUBJECT'S LAST NAME: HYDE

DATE/TIME OF TEST: 09/04/2016 19:37:44

Diagnostics Check		OK	19:31
Air Blank		0.000	19:31
Control Test		0.077	19:32
Air Blank		0.000	19:32
Subject Sample	# 1	0.204	19:33
Breath Volume		4.062	19:33
Air Blank		0.000	19:33
Air Blank		0.000	19:35
Subject Sample	# 2	0.200	19:36
Breath Volume		3.097	19:36
Air Blank		0.000	19:36
Control Test		0.076	19:37
Air Blank		0.000	19:37
Diagnostics Check		OK	19:37

Data Download Date/Time: 09/06/2016 12:06

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOUD,JOHN,D
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: CHRISTELLO,CLIFFORD
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 08/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:21

SUBJECT'S LAST NAME: SELLNER

DATE/TIME OF TEST: 09/06/2016 03:03:06

Diagnostics Check		OK	02:56
Air Blank		0.000	02:56
Control Test		0.076	02:56
Air Blank		0.000	02:57
Subject Sample	# 1	0.216	02:58
Breath Volume		1.558	02:58
Air Blank		0.000	02:59
Air Blank		0.000	03:00
Subject Sample	# 2	0.216	03:01
Breath Volume		1.488	03:01
Air Blank		0.000	03:02
Control Test		0.075	03:02
Air Blank		0.000	03:03
Diagnostics Check		OK	03:03

Data Download Date/Time: 09/06/2016 12:06

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001056

INSTRUMENT REGISTERED TO: DAVIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAINES,DAVID,K
BREATH TEST OPERATOR AGENCY: Davie PD
ARRESTING OFFICER: YOUNG,JOSHUA
ARRESTING OFFICER AGENCY: Davie PD
LAST AGENCY INSPECTION: 08/31/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG429701
DRY GAS EXPIRY DATE: 10/24/2016
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:19

SUBJECT'S LAST NAME: KISSOONDATH

DATE/TIME OF TEST: 09/04/2016 01:52:24

Diagnostics Check		OK	01:45
Air Blank		0.000	01:46
Control Test		0.079	01:46
Air Blank		0.000	01:46
Subject Sample	# 1	0.168	01:47
Breath Volume		2.671	01:47
Air Blank		0.000	01:48
Air Blank		0.000	01:50
Subject Sample	# 2	0.158	01:50
Breath Volume		2.691	01:50
Air Blank		0.000	01:51
Control Test		0.080	01:51
Air Blank		0.000	01:52
Diagnostics Check		OK	01:52

Data Download Date/Time: 09/30/2016 16:02

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001056

INSTRUMENT REGISTERED TO: DAVIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DI CINTIO,PATRICK,N
BREATH TEST OPERATOR AGENCY: Davie PD
ARRESTING OFFICER: YOUNG,JOSHUA
ARRESTING OFFICER AGENCY: Davie PD
LAST AGENCY INSPECTION: 08/31/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG429701
DRY GAS EXPIRY DATE: 10/24/2016
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:11

SUBJECT'S LAST NAME: GELBMANN

DATE/TIME OF TEST: 09/18/2016 01:53:00

Diagnostics Check		OK	01:46
Air Blank		0.000	01:47
Control Test		0.075	01:47
Air Blank		0.000	01:47
Subject Sample	# 1	0.282	01:48
Breath Volume		3.296	01:48
Air Blank		0.000	01:49
Air Blank		0.000	01:50
Subject Sample	# 2	0.268	01:51
Breath Volume		3.738	01:51
Air Blank		0.000	01:52
Control Test		0.077	01:52
Air Blank		0.000	01:52
Diagnostics Check		OK	01:53

Data Download Date/Time: 09/30/2016 16:02

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001059

INSTRUMENT REGISTERED TO: FERNANDINA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,JORGE,L
BREATH TEST OPERATOR AGENCY: Fernandina Beach PD
ARRESTING OFFICER: HERNANDEZ,JORGE
ARRESTING OFFICER AGENCY: Fernandina Beach PD
LAST AGENCY INSPECTION: 09/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A3
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:04

SUBJECT'S LAST NAME: ARIAS

DATE/TIME OF TEST: 09/14/2016 01:37:22

Diagnostics Check		OK	01:31
Air Blank		0.000	01:31
Control Test		0.078	01:31
Air Blank		0.000	01:32
Subject Sample	# 1	0.335	01:32
Breath Volume		2.175	01:32
Air Blank		0.000	01:33
Air Blank		0.000	01:35
Subject Sample	# 2	0.331	01:35
Breath Volume		1.640	01:35
Air Blank		0.000	01:36
Control Test		0.079	01:36
Air Blank		0.000	01:37
Diagnostics Check		OK	01:37

Data Download Date/Time: 10/12/2016 15:42

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001059

INSTRUMENT REGISTERED TO: FERNANDINA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAETSCH,JOSH,E
BREATH TEST OPERATOR AGENCY: Fernandina Beach PD
ARRESTING OFFICER: PAETSCH,PAETSCH
ARRESTING OFFICER AGENCY: Fernandina Beach PD
LAST AGENCY INSPECTION: 09/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A3
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:09

SUBJECT'S LAST NAME: MURRAY

DATE/TIME OF TEST: 09/23/2016 03:41:48

Diagnostics Check		OK	03:38	
Air Blank		0.000	03:39	
Control Test		0.078	03:39	
Air Blank		0.000	03:40	
Subject Sample	# 1		03:40	Subject Test Refused
Breath Volume		0.000	03:40	
Air Blank		0.000	03:40	
Control Test		0.078	03:41	
Air Blank		0.000	03:41	
Diagnostics Check		OK	03:41	

Data Download Date/Time: 10/12/2016 15:42

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001060

INSTRUMENT REGISTERED TO: WINTER HAVEN PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NGUYEN,JOHN,
BREATH TEST OPERATOR AGENCY: Winter Haven PD
ARRESTING OFFICER: LEWIS,RYAN
ARRESTING OFFICER AGENCY: Winter Haven PD
LAST AGENCY INSPECTION: 08/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:16

SUBJECT'S LAST NAME: ROSEN

DATE/TIME OF TEST: 09/24/2016 00:45:14

Diagnostics Check		OK	00:38
Air Blank		0.000	00:38
Control Test		0.081	00:39
Air Blank		0.000	00:39
Subject Sample	# 1	0.191	00:40
Breath Volume		2.710	00:40
Air Blank		0.000	00:41
Air Blank		0.000	00:42
Subject Sample	# 2	0.190	00:43
Breath Volume		2.503	00:43
Air Blank		0.000	00:44
Control Test		0.080	00:44
Air Blank		0.000	00:45
Diagnostics Check		OK	00:45

Data Download Date/Time: 09/29/2016 08:52

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001061

INSTRUMENT REGISTERED TO: GULFPORT PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAROTTA,MICHAEL,S
BREATH TEST OPERATOR AGENCY: Gulfport PD
ARRESTING OFFICER: CARTER,TIMOTHY
ARRESTING OFFICER AGENCY: Gulfport PD
LAST AGENCY INSPECTION: 08/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A3
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:34

SUBJECT'S LAST NAME: SMALLS

DATE/TIME OF TEST: 09/10/2016 04:00:50

Diagnostics Check		OK	03:54
Air Blank		0.000	03:55
Control Test		0.080	03:55
Air Blank		0.000	03:55
Subject Sample	# 1	0.198	03:56
Breath Volume		1.265	03:56
Air Blank		0.000	03:57
Air Blank		0.000	03:58
Subject Sample	# 2	0.200	03:59
Breath Volume		1.367	03:59
Air Blank		0.000	03:59
Control Test		0.080	04:00
Air Blank		0.000	04:00
Diagnostics Check		OK	04:00

Data Download Date/Time: 09/30/2016 03:14

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001061

INSTRUMENT REGISTERED TO: GULFPORT PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAROTTA,MICHAEL,S
BREATH TEST OPERATOR AGENCY: Gulfport PD
ARRESTING OFFICER: CARTER,TIMOTHY
ARRESTING OFFICER AGENCY: Gulfport PD
LAST AGENCY INSPECTION: 08/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A3
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: HOFFMAN

DATE/TIME OF TEST: 09/15/2016 01:03:47

Diagnostics Check		OK	00:51	
Air Blank		0.000	00:51	
Control Test		0.079	00:52	
Air Blank		0.000	00:52	
Subject Sample	# 1	0.293	00:55	Volume Not Met
Breath Volume		0.273	00:55	
Air Blank		0.000	00:56	
Air Blank		0.000	00:58	
Subject Sample	# 2	0.337	00:58	
Breath Volume		1.367	00:58	
Air Blank		0.000	00:59	
Air Blank		0.000	01:01	
Subject Sample	# 3	0.356	01:02	
Breath Volume		1.625	01:02	
Air Blank		0.000	01:02	
Control Test		0.079	01:03	
Air Blank		0.000	01:03	
Diagnostics Check		OK	01:03	

Data Download Date/Time: 09/30/2016 03:14

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001062

INSTRUMENT REGISTERED TO: TREASURE ISLAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DORMAN,ZAKARIAH,S
BREATH TEST OPERATOR AGENCY: Treasure Island PD
ARRESTING OFFICER: MORTON,DAN
ARRESTING OFFICER AGENCY: Treasure Island PD
LAST AGENCY INSPECTION: 09/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A6
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 16:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:10

SUBJECT'S LAST NAME: PRIMAVERA

DATE/TIME OF TEST: 09/08/2016 17:57:05

Diagnostics Check		OK	17:50
Air Blank		0.000	17:50
Control Test		0.080	17:51
Air Blank		0.000	17:51
Subject Sample	# 1	0.210	17:52
Breath Volume		1.355	17:52
Air Blank		0.000	17:53
Air Blank		0.000	17:55
Subject Sample	# 2	0.209	17:55
Breath Volume		1.800	17:55
Air Blank		0.000	17:56
Control Test		0.080	17:56
Air Blank		0.000	17:57
Diagnostics Check		OK	17:57

Data Download Date/Time: 10/07/2016 12:03

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001062

INSTRUMENT REGISTERED TO: TREASURE ISLAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CONTOIS,R,
BREATH TEST OPERATOR AGENCY: Treasure Island PD
ARRESTING OFFICER: CONTOIS,R
ARRESTING OFFICER AGENCY: Treasure Island PD
LAST AGENCY INSPECTION: 09/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A6
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:28

SUBJECT'S LAST NAME: ELSASSER

DATE/TIME OF TEST: 09/11/2016 22:57:33

Diagnostics Check		OK	22:54	
Air Blank		0.000	22:55	
Control Test		0.081	22:55	
Air Blank		0.000	22:55	
Subject Sample	# 1		22:56	Subject Test Refused
Breath Volume		0.000	22:56	
Air Blank		0.000	22:56	
Control Test		0.081	22:56	
Air Blank		0.000	22:57	
Diagnostics Check		OK	22:57	

Data Download Date/Time: 10/07/2016 12:03

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001062

INSTRUMENT REGISTERED TO: TREASURE ISLAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CONTOIS,R,
BREATH TEST OPERATOR AGENCY: Treasure Island PD
ARRESTING OFFICER: CONTOIS,R
ARRESTING OFFICER AGENCY: Treasure Island PD
LAST AGENCY INSPECTION: 09/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A6
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:54

SUBJECT'S LAST NAME: A352470852230

DATE/TIME OF TEST: 09/13/2016 03:31:48

Diagnostics Check		OK	03:27	
Air Blank		0.000	03:28	
Control Test		0.080	03:28	
Air Blank		0.000	03:29	
Subject Sample	# 1		03:30	Subject Test Refused
Breath Volume		0.000	03:30	
Air Blank		0.000	03:30	
Control Test		0.080	03:31	
Air Blank		0.000	03:31	
Diagnostics Check		OK	03:31	

Data Download Date/Time: 10/07/2016 12:03

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001062

INSTRUMENT REGISTERED TO: TREASURE ISLAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: INNES,JOSHUA,L
BREATH TEST OPERATOR AGENCY: Treasure Island PD
ARRESTING OFFICER: INNES,JOSHUA
ARRESTING OFFICER AGENCY: Treasure Island PD
LAST AGENCY INSPECTION: 09/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A6
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 16:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:58

SUBJECT'S LAST NAME: NICKERSON

DATE/TIME OF TEST: 09/17/2016 17:26:30

Diagnostics Check		OK	17:23	
Air Blank		0.000	17:24	
Control Test		0.081	17:24	
Air Blank		0.000	17:24	
Subject Sample	# 1		17:25	Subject Test Refused
Breath Volume		0.000	17:25	
Air Blank		0.000	17:25	
Control Test		0.081	17:25	
Air Blank		0.000	17:26	
Diagnostics Check		OK	17:26	

Data Download Date/Time: 10/07/2016 12:03

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

INSTRUMENT REGISTERED TO: NEW PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ECK,AARON,R
BREATH TEST OPERATOR AGENCY: New Port Richey PD
ARRESTING OFFICER: ROWAND,ART
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 08/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29614080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 01:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:02

SUBJECT'S LAST NAME: WATKINS

DATE/TIME OF TEST: 09/08/2016 02:29:57

Diagnostics Check		OK	02:23
Air Blank		0.000	02:23
Control Test		0.076	02:24
Air Blank		0.000	02:24
Subject Sample	# 1	0.278	02:25
Breath Volume		2.453	02:25
Air Blank		0.000	02:26
Air Blank		0.000	02:27
Subject Sample	# 2	0.293	02:28
Breath Volume		3.042	02:28
Air Blank		0.000	02:29
Control Test		0.077	02:29
Air Blank		0.000	02:29
Diagnostics Check		OK	02:29

Data Download Date/Time: 09/14/2016 21:22

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001065

INSTRUMENT REGISTERED TO: TARPON SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEMMON,CHRIS,F
BREATH TEST OPERATOR AGENCY: Tarpon Springs PD
ARRESTING OFFICER: LEMMON,CHRIS
ARRESTING OFFICER AGENCY: Tarpon Springs PD
LAST AGENCY INSPECTION: 08/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A2
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:38

SUBJECT'S LAST NAME: WORDEN

DATE/TIME OF TEST: 09/03/2016 03:02:27

Diagnostics Check		OK	02:59	
Air Blank		0.000	03:00	
Control Test		0.077	03:00	
Air Blank		0.000	03:00	
Subject Sample	# 1		03:01	Subject Test Refused
Breath Volume		0.000	03:01	
Air Blank		0.000	03:01	
Control Test		0.078	03:01	
Air Blank		0.000	03:02	
Diagnostics Check		OK	03:02	

Data Download Date/Time: 09/30/2016 21:35

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001065

INSTRUMENT REGISTERED TO: TARPON SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIBSON,JOHN,A
BREATH TEST OPERATOR AGENCY: Tarpon Springs PD
ARRESTING OFFICER: GIBSON,JOHN
ARRESTING OFFICER AGENCY: Tarpon Springs PD
LAST AGENCY INSPECTION: 08/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A2
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:38

SUBJECT'S LAST NAME: PIMENTEL FERREIRA

DATE/TIME OF TEST: 09/03/2016 05:07:55

Diagnostics Check		OK	05:00
Air Blank		0.000	05:01
Control Test		0.077	05:01
Air Blank		0.000	05:01
Subject Sample	# 1	0.072	05:02
Breath Volume		1.136	05:02
Air Blank		0.000	05:03
Air Blank		0.000	05:05
Subject Sample	# 2	0.073	05:06
Breath Volume		2.207	05:06
Air Blank		0.000	05:07
Control Test		0.077	05:07
Air Blank		0.000	05:07
Diagnostics Check		OK	05:07

Data Download Date/Time: 09/30/2016 21:36

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001065

INSTRUMENT REGISTERED TO: TARPON SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIBSON,JOHN,A
BREATH TEST OPERATOR AGENCY: Tarpon Springs PD
ARRESTING OFFICER: HOLLINGSWORTH,JACOB
ARRESTING OFFICER AGENCY: Tarpon Springs PD
LAST AGENCY INSPECTION: 08/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A2
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:32

SUBJECT'S LAST NAME: AUTEN

DATE/TIME OF TEST: 09/13/2016 05:29:51

Diagnostics Check		OK	05:23
Air Blank		0.000	05:24
Control Test		0.077	05:24
Air Blank		0.000	05:24
Subject Sample	# 1	0.052	05:25
Breath Volume		4.015	05:25
Air Blank		0.000	05:25
Air Blank		0.000	05:27
Subject Sample	# 2	0.050	05:28
Breath Volume		3.437	05:28
Air Blank		0.000	05:28
Control Test		0.078	05:29
Air Blank		0.000	05:29
Diagnostics Check		OK	05:29

Data Download Date/Time: 09/30/2016 21:36

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAIR,RONALD,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: JASIUKIEWICZ,MICHAEL
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 08/27/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 21:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:35

SUBJECT'S LAST NAME: PRODROMITIS

DATE/TIME OF TEST: 09/02/2016 22:01:40

Diagnostics Check		OK	21:55
Air Blank		0.000	21:55
Control Test		0.077	21:55
Air Blank		0.000	21:56
Subject Sample	# 1	0.152	21:57
Breath Volume		2.164	21:57
Air Blank		0.000	21:57
Air Blank		0.000	21:59
Subject Sample	# 2	0.142	22:00
Breath Volume		1.375	22:00
Air Blank		0.000	22:00
Control Test		0.077	22:01
Air Blank		0.000	22:01
Diagnostics Check		OK	22:01

Data Download Date/Time: 09/29/2016 21:06

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAIR,RONALD,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LANEY,DAMON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 08/27/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 23:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:37

SUBJECT'S LAST NAME: COOK

DATE/TIME OF TEST: 09/03/2016 00:00:12

Diagnostics Check		OK	23:57	
Air Blank		0.000	23:57	
Control Test		0.077	23:58	
Air Blank		0.000	23:58	
Subject Sample	# 1		23:58	Subject Test Refused
Breath Volume		0.000	23:58	
Air Blank		0.000	23:59	
Control Test		0.077	23:59	
Air Blank		0.000	00:00	
Diagnostics Check		OK	00:00	

Data Download Date/Time: 09/29/2016 21:06

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAIR,RONALD,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WAGNER,WAYNE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 08/27/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 00:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:42

SUBJECT'S LAST NAME: LESSARD

DATE/TIME OF TEST: 09/03/2016 01:05:17

Diagnostics Check		OK	01:02	
Air Blank		0.000	01:02	
Control Test		0.078	01:03	
Air Blank		0.000	01:03	
Subject Sample	# 1		01:03	Subject Test Refused
Breath Volume		0.000	01:03	
Air Blank		0.000	01:04	
Control Test		0.077	01:04	
Air Blank		0.000	01:05	
Diagnostics Check		OK	01:05	

Data Download Date/Time: 09/29/2016 21:06

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DONNELLY,JONATHAN,P
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: DONNELLY,JONATHAN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 08/27/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 01:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: JAMES

DATE/TIME OF TEST: 09/04/2016 02:52:22

Diagnostics Check		OK	02:45
Air Blank		0.000	02:46
Control Test		0.079	02:46
Air Blank		0.000	02:47
Subject Sample	# 1	0.097	02:47
Breath Volume		2.058	02:47
Air Blank		0.000	02:48
Air Blank		0.000	02:50
Subject Sample	# 2	0.104	02:50
Breath Volume		1.453	02:50
Air Blank		0.000	02:51
Control Test		0.079	02:51
Air Blank		0.000	02:52
Diagnostics Check		OK	02:52

Data Download Date/Time: 09/29/2016 21:06

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCKENZIE,DEPUTY,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MCKENZIE,DEPUTY
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 08/27/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 15:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:50

SUBJECT'S LAST NAME: LABUS

DATE/TIME OF TEST: 09/06/2016 17:13:58

Diagnostics Check		OK	17:10	
Air Blank		0.000	17:11	
Control Test		0.079	17:11	
Air Blank		0.000	17:12	
Subject Sample	# 1		17:12	Interferent Detect
Breath Volume		2.000	17:12	
Air Blank		0.000	17:13	
Control Test		0.079	17:13	
Air Blank		0.000	17:13	
Diagnostics Check		OK	17:13	

Data Download Date/Time: 09/29/2016 21:06

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DONNELLY,JONATHAN,P
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: DONNELLY,JONATHAN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 08/27/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 01:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:29

SUBJECT'S LAST NAME: BAKSIS

DATE/TIME OF TEST: 09/11/2016 02:56:03

Diagnostics Check		OK	02:49
Air Blank		0.000	02:50
Control Test		0.079	02:50
Air Blank		0.000	02:51
Subject Sample	# 1	0.160	02:51
Breath Volume		1.406	02:51
Air Blank		0.000	02:52
Air Blank		0.000	02:54
Subject Sample	# 2	0.155	02:54
Breath Volume		1.425	02:54
Air Blank		0.000	02:55
Control Test		0.078	02:55
Air Blank		0.000	02:55
Diagnostics Check		OK	02:56

Data Download Date/Time: 09/29/2016 21:06

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLE,S,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BAGLI,M
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 08/27/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 01:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: BICKETT

DATE/TIME OF TEST: 09/19/2016 03:01:07

Diagnostics Check		OK	02:55
Air Blank		0.000	02:55
Control Test		0.080	02:55
Air Blank		0.000	02:56
Subject Sample	# 1	0.175	02:56
Breath Volume		2.226	02:56
Air Blank		0.000	02:57
Air Blank		0.000	02:59
Subject Sample	# 2	0.175	02:59
Breath Volume		1.750	02:59
Air Blank		0.000	03:00
Control Test		0.079	03:00
Air Blank		0.000	03:01
Diagnostics Check		OK	03:01

Data Download Date/Time: 09/29/2016 21:06

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: FERGUSON,B
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 08/27/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 03:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: TRUSKEY

DATE/TIME OF TEST: 09/29/2016 04:20:41

Diagnostics Check		OK	04:12
Air Blank		0.000	04:13
Control Test		0.080	04:13
Air Blank		0.000	04:13
Subject Sample	# 1	0.117	04:15
Breath Volume		2.179	04:15
Air Blank		0.000	04:16
Air Blank		0.000	04:18
Subject Sample	# 2	0.133	04:19
Breath Volume		1.769	04:19
Air Blank		0.000	04:19
Control Test		0.079	04:20
Air Blank		0.000	04:20
Diagnostics Check		OK	04:20

Data Download Date/Time: 09/29/2016 21:06

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARNANDAN,INDROWTIE,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: THOMPSON,BRETT
ARRESTING OFFICER AGENCY: Pembroke Pines PD
LAST AGENCY INSPECTION: 08/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 00:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:12

SUBJECT'S LAST NAME: CHENG-HSIANG

DATE/TIME OF TEST: 09/01/2016 01:44:12

Diagnostics Check		OK	01:32	
Air Blank		0.000	01:33	
Control Test		0.079	01:33	
Air Blank		0.000	01:33	
Subject Sample	# 1	0.060	01:37	Volume Not Met
Breath Volume		0.070	01:37	
Air Blank		0.000	01:37	
Air Blank		0.000	01:39	
Subject Sample	# 2	0.087	01:42	Volume Not Met
Breath Volume		0.019	01:42	
Air Blank		0.000	01:43	
Control Test		0.079	01:43	
Air Blank		0.000	01:44	
Diagnostics Check		OK	01:44	

Data Download Date/Time: 09/24/2016 20:48

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARNANDAN,INDROWTIE,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: MANNING,MARC
ARRESTING OFFICER AGENCY: Pembroke Pines PD
LAST AGENCY INSPECTION: 08/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 04:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:50

SUBJECT'S LAST NAME: MILLINE

DATE/TIME OF TEST: 09/02/2016 05:18:16

Diagnostics Check		OK	05:10
Air Blank		0.000	05:11
Control Test		0.078	05:11
Air Blank		0.000	05:12
Subject Sample	# 1	0.178	05:13
Breath Volume		1.640	05:13
Air Blank		0.000	05:13
Air Blank		0.000	05:15
Subject Sample	# 2	0.163	05:16
Breath Volume		1.757	05:16
Air Blank		0.000	05:17
Control Test		0.078	05:17
Air Blank		0.000	05:18
Diagnostics Check		OK	05:18

Data Download Date/Time: 09/24/2016 20:48

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARNANDAN,INDROWTIE,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: DUSENBERY,ASHLEY
ARRESTING OFFICER AGENCY: Sunrise PD
LAST AGENCY INSPECTION: 08/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:26

SUBJECT'S LAST NAME: KESSLER

DATE/TIME OF TEST: 09/03/2016 01:52:50

Diagnostics Check		OK	01:46
Air Blank		0.000	01:47
Control Test		0.079	01:47
Air Blank		0.000	01:48
Subject Sample	# 1	0.170	01:48
Breath Volume		2.949	01:48
Air Blank		0.000	01:49
Air Blank		0.000	01:50
Subject Sample	# 2	0.163	01:51
Breath Volume		2.269	01:51
Air Blank		0.000	01:51
Control Test		0.079	01:52
Air Blank		0.000	01:52
Diagnostics Check		OK	01:52

Data Download Date/Time: 09/24/2016 20:48

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARNANDAN,INDROWTIE,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SPEZIALE,ADAM
ARRESTING OFFICER AGENCY: Margate PD
LAST AGENCY INSPECTION: 08/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 03:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: REY

DATE/TIME OF TEST: 09/03/2016 04:01:19

Diagnostics Check		OK	03:55
Air Blank		0.000	03:55
Control Test		0.079	03:56
Air Blank		0.000	03:56
Subject Sample	# 1	0.170	03:56
Breath Volume		1.371	03:56
Air Blank		0.000	03:57
Air Blank		0.000	03:59
Subject Sample	# 2	0.168	03:59
Breath Volume		1.820	03:59
Air Blank		0.000	04:00
Control Test		0.079	04:00
Air Blank		0.000	04:01
Diagnostics Check		OK	04:01

Data Download Date/Time: 09/24/2016 20:48

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SAPP,JOSHUA
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 08/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 00:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:57

SUBJECT'S LAST NAME: BAPTISTE

DATE/TIME OF TEST: 09/04/2016 01:24:17

Diagnostics Check		OK	01:17
Air Blank		0.000	01:17
Control Test		0.079	01:18
Air Blank		0.000	01:18
Subject Sample	# 1	0.180	01:19
Breath Volume		1.246	01:19
Air Blank		0.000	01:20
Air Blank		0.000	01:22
Subject Sample	# 2	0.187	01:22
Breath Volume		1.453	01:22
Air Blank		0.000	01:23
Control Test		0.079	01:23
Air Blank		0.000	01:24
Diagnostics Check		OK	01:24

Data Download Date/Time: 09/24/2016 20:48

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: POPICK,SCOTT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 08/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 00:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:42

SUBJECT'S LAST NAME: NEEDHAM

DATE/TIME OF TEST: 09/04/2016 02:12:02

Diagnostics Check		OK	02:04
Air Blank		0.000	02:05
Control Test		0.079	02:05
Air Blank		0.000	02:05
Subject Sample	# 1	0.180	02:07
Breath Volume		2.976	02:07
Air Blank		0.000	02:07
Air Blank		0.000	02:09
Subject Sample	# 2	0.177	02:10
Breath Volume		1.878	02:10
Air Blank		0.000	02:11
Control Test		0.078	02:11
Air Blank		0.000	02:11
Diagnostics Check		OK	02:12

Data Download Date/Time: 09/24/2016 20:48

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SERPHOS,STEVEN
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 08/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:32

SUBJECT'S LAST NAME: DE NOBREGA FOJ

DATE/TIME OF TEST: 09/04/2016 03:03:45

Diagnostics Check		OK	02:56
Air Blank		0.000	02:57
Control Test		0.080	02:57
Air Blank		0.000	02:58
Subject Sample	# 1	0.105	02:59
Breath Volume		1.222	02:59
Air Blank		0.000	02:59
Air Blank		0.000	03:01
Subject Sample	# 2	0.100	03:02
Breath Volume		1.273	03:02
Air Blank		0.000	03:02
Control Test		0.079	03:03
Air Blank		0.000	03:03
Diagnostics Check		OK	03:03

Data Download Date/Time: 09/24/2016 20:48

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: DEDEJ,MEDJON
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 08/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 02:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:23

SUBJECT'S LAST NAME: FUNK

DATE/TIME OF TEST: 09/06/2016 03:52:29

Diagnostics Check		OK	03:43
Air Blank		0.000	03:44
Control Test		0.080	03:44
Air Blank		0.000	03:45
Subject Sample	# 1	0.204	03:46
Breath Volume		1.628	03:46
Air Blank		0.000	03:46
Air Blank		0.000	03:48
Subject Sample	# 2	0.192	03:50
Breath Volume		1.550	03:50
Air Blank		0.000	03:51
Control Test		0.079	03:51
Air Blank		0.000	03:52
Diagnostics Check		OK	03:52

Data Download Date/Time: 09/24/2016 20:48

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: DERIET,RAFAEL
ARRESTING OFFICER AGENCY: Hollywood PD
LAST AGENCY INSPECTION: 08/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 01:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:04

SUBJECT'S LAST NAME: GAONA GRANADA

DATE/TIME OF TEST: 09/10/2016 02:33:45

Diagnostics Check		OK	02:27
Air Blank		0.000	02:27
Control Test		0.080	02:27
Air Blank		0.000	02:28
Subject Sample	# 1	0.223	02:29
Breath Volume		3.117	02:29
Air Blank		0.000	02:29
Air Blank		0.000	02:31
Subject Sample	# 2	0.220	02:32
Breath Volume		3.000	02:32
Air Blank		0.000	02:32
Control Test		0.078	02:33
Air Blank		0.000	02:33
Diagnostics Check		OK	02:33

Data Download Date/Time: 09/24/2016 20:48

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: PARRISH,MATTHEW
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 08/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 07:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:19

SUBJECT'S LAST NAME: GORDON

DATE/TIME OF TEST: 09/10/2016 09:54:25

Diagnostics Check		OK	09:48
Air Blank		0.000	09:48
Control Test		0.080	09:48
Air Blank		0.000	09:49
Subject Sample	# 1	0.198	09:49
Breath Volume		2.042	09:49
Air Blank		0.000	09:50
Air Blank		0.000	09:52
Subject Sample	# 2	0.201	09:52
Breath Volume		1.835	09:52
Air Blank		0.000	09:53
Control Test		0.080	09:53
Air Blank		0.000	09:54
Diagnostics Check		OK	09:54

Data Download Date/Time: 09/24/2016 20:48

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUBER,GERALD,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: CAROTTI,MOISES
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 08/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 22:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:17

SUBJECT'S LAST NAME: HUBER

DATE/TIME OF TEST: 09/10/2016 23:50:44

Diagnostics Check		OK	23:44
Air Blank		0.000	23:44
Control Test		0.079	23:44
Air Blank		0.000	23:45
Subject Sample	# 1	0.199	23:46
Breath Volume		1.820	23:46
Air Blank		0.000	23:46
Air Blank		0.000	23:48
Subject Sample	# 2	0.198	23:49
Breath Volume		1.433	23:49
Air Blank		0.000	23:49
Control Test		0.079	23:50
Air Blank		0.000	23:50
Diagnostics Check		OK	23:50

Data Download Date/Time: 09/24/2016 20:48

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINA,MARMILLY,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: STURM,FREDRIC
ARRESTING OFFICER AGENCY: Margate PD
LAST AGENCY INSPECTION: 08/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 03:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:25

SUBJECT'S LAST NAME: AREUETA

DATE/TIME OF TEST: 09/11/2016 05:02:53

Diagnostics Check		OK	04:56
Air Blank		0.000	04:56
Control Test		0.079	04:57
Air Blank		0.000	04:57
Subject Sample	# 1	0.167	04:58
Breath Volume		2.218	04:58
Air Blank		0.000	04:59
Air Blank		0.000	05:00
Subject Sample	# 2	0.167	05:01
Breath Volume		1.183	05:01
Air Blank		0.000	05:01
Control Test		0.080	05:02
Air Blank		0.000	05:02
Diagnostics Check		OK	05:02

Data Download Date/Time: 09/24/2016 20:48

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINA,MARMILLY,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GRECO,SAL
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 08/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 11:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:10

SUBJECT'S LAST NAME: INFANTE

DATE/TIME OF TEST: 09/11/2016 13:49:13

Diagnostics Check		OK	13:42
Air Blank		0.000	13:43
Control Test		0.080	13:43
Air Blank		0.000	13:44
Subject Sample	# 1	0.000	13:44
Breath Volume		2.558	13:44
Air Blank		0.000	13:45
Air Blank		0.000	13:47
Subject Sample	# 2	0.000	13:47
Breath Volume		1.652	13:47
Air Blank		0.000	13:48
Control Test		0.080	13:48
Air Blank		0.000	13:49
Diagnostics Check		OK	13:49

Data Download Date/Time: 09/24/2016 20:48

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ATTIKISSON,MELVIN
ARRESTING OFFICER AGENCY: Hollywood PD
LAST AGENCY INSPECTION: 08/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 00:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:14

SUBJECT'S LAST NAME: MONTALVO

DATE/TIME OF TEST: 09/12/2016 01:42:30

Diagnostics Check		OK	01:36
Air Blank		0.000	01:37
Control Test		0.080	01:37
Air Blank		0.000	01:37
Subject Sample	# 1	0.000	01:38
Breath Volume		1.968	01:38
Air Blank		0.000	01:38
Air Blank		0.000	01:40
Subject Sample	# 2	0.000	01:41
Breath Volume		3.214	01:41
Air Blank		0.000	01:41
Control Test		0.079	01:41
Air Blank		0.000	01:42
Diagnostics Check		OK	01:42

Data Download Date/Time: 09/24/2016 20:48

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: BLACK,GORDON
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 08/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 18:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:39

SUBJECT'S LAST NAME: COGHILL JR

DATE/TIME OF TEST: 09/12/2016 19:08:58

Diagnostics Check		OK	19:02
Air Blank		0.000	19:03
Control Test		0.079	19:03
Air Blank		0.000	19:03
Subject Sample	# 1	0.181	19:04
Breath Volume		2.222	19:04
Air Blank		0.000	19:05
Air Blank		0.000	19:06
Subject Sample	# 2	0.177	19:07
Breath Volume		2.007	19:07
Air Blank		0.000	19:08
Control Test		0.079	19:08
Air Blank		0.000	19:08
Diagnostics Check		OK	19:08

Data Download Date/Time: 09/24/2016 20:48

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARNANDAN,INDROWTIE,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SERPHOS,STEVEN
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 08/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 01:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:42

SUBJECT'S LAST NAME: MEJIA SOTO

DATE/TIME OF TEST: 09/16/2016 02:09:28

Diagnostics Check		OK	02:02
Air Blank		0.000	02:02
Control Test		0.079	02:03
Air Blank		0.000	02:03
Subject Sample	# 1	0.069	02:04
Breath Volume		1.667	02:04
Air Blank		0.000	02:05
Air Blank		0.000	02:07
Subject Sample	# 2	0.073	02:08
Breath Volume		1.667	02:08
Air Blank		0.000	02:08
Control Test		0.079	02:08
Air Blank		0.000	02:09
Diagnostics Check		OK	02:09

Data Download Date/Time: 09/24/2016 20:48

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARNANDAN,INDROWTIE,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: WILLIAMS,SEAN
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 08/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 03:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:58

SUBJECT'S LAST NAME: BAUGH

DATE/TIME OF TEST: 09/16/2016 04:24:27

Diagnostics Check		OK	04:18
Air Blank		0.000	04:19
Control Test		0.080	04:19
Air Blank		0.000	04:19
Subject Sample	# 1	0.000	04:20
Breath Volume		2.550	04:20
Air Blank		0.000	04:20
Air Blank		0.000	04:22
Subject Sample	# 2	0.000	04:23
Breath Volume		2.636	04:23
Air Blank		0.000	04:23
Control Test		0.079	04:23
Air Blank		0.000	04:24
Diagnostics Check		OK	04:24

Data Download Date/Time: 09/24/2016 20:48

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARNANDAN,INDROWTIE,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: DOMINGUEZ,ASIEL
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 08/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 00:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 09/17/2016 01:10:50

Diagnostics Check		OK	01:04
Air Blank		0.000	01:04
Control Test		0.080	01:05
Air Blank		0.000	01:05
Subject Sample	# 1	0.188	01:06
Breath Volume		2.851	01:06
Air Blank		0.000	01:06
Air Blank		0.000	01:08
Subject Sample	# 2	0.191	01:09
Breath Volume		2.269	01:09
Air Blank		0.000	01:09
Control Test		0.079	01:10
Air Blank		0.000	01:10
Diagnostics Check		OK	01:10

Data Download Date/Time: 09/24/2016 20:48

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: WASHINGTON,YVETTE
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 08/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 02:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:49

SUBJECT'S LAST NAME: IZQUIERDO

DATE/TIME OF TEST: 09/17/2016 03:17:26

Diagnostics Check		OK	03:11
Air Blank		0.000	03:11
Control Test		0.080	03:12
Air Blank		0.000	03:12
Subject Sample	# 1	0.116	03:12
Breath Volume		1.214	03:12
Air Blank		0.000	03:13
Air Blank		0.000	03:15
Subject Sample	# 2	0.114	03:15
Breath Volume		1.433	03:15
Air Blank		0.000	03:16
Control Test		0.079	03:16
Air Blank		0.000	03:17
Diagnostics Check		OK	03:17

Data Download Date/Time: 09/24/2016 20:48

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARNANDAN,INDROWTIE,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: CARELLI,MICHAEL
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 08/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 05:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:40

SUBJECT'S LAST NAME: CASTILLO CASTILLO

DATE/TIME OF TEST: 09/17/2016 07:06:59

Diagnostics Check		OK	07:00
Air Blank		0.000	07:00
Control Test		0.080	07:00
Air Blank		0.000	07:01
Subject Sample	# 1	0.145	07:02
Breath Volume		1.203	07:02
Air Blank		0.000	07:02
Air Blank		0.000	07:04
Subject Sample	# 2	0.159	07:05
Breath Volume		2.156	07:05
Air Blank		0.000	07:06
Control Test		0.079	07:06
Air Blank		0.000	07:06
Diagnostics Check		OK	07:06

Data Download Date/Time: 09/24/2016 20:48

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SAPP,JOSHUA
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 08/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 02:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: GLOVER

DATE/TIME OF TEST: 09/18/2016 03:58:26

Diagnostics Check		OK	03:45	
Air Blank		0.000	03:45	
Control Test		0.079	03:46	
Air Blank		0.000	03:46	
Subject Sample	# 1	0.086	03:49	Volume Not Met
Breath Volume		0.246	03:49	
Air Blank		0.000	03:50	
Air Blank		0.000	03:52	
Subject Sample	# 2	0.085	03:53	
Breath Volume		2.945	03:53	
Air Blank		0.000	03:54	
Air Blank		0.000	03:56	
Subject Sample	# 3	0.080	03:56	
Breath Volume		2.792	03:56	
Air Blank		0.000	03:57	
Control Test		0.079	03:57	
Air Blank		0.000	03:58	
Diagnostics Check		OK	03:58	

Data Download Date/Time: 09/24/2016 20:48

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SERPHOS,STEVEN
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 08/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 03:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:24

SUBJECT'S LAST NAME: OTOYA

DATE/TIME OF TEST: 09/18/2016 04:24:05

Diagnostics Check		OK	04:09	
Air Blank		0.000	04:10	
Control Test		0.078	04:10	
Air Blank		0.000	04:11	
Subject Sample	# 1	0.182	04:14	
Breath Volume		1.382	04:14	
Air Blank		0.000	04:14	
Air Blank		0.000	04:16	
Subject Sample	# 2	0.174	04:19	Volume Not Met
Breath Volume		0.441	04:19	
Air Blank		0.000	04:20	
Air Blank		0.000	04:22	
Subject Sample	# 3		04:22	Subject Test Refused
Breath Volume		0.000	04:22	
Air Blank		0.000	04:23	
Control Test		0.078	04:23	
Air Blank		0.000	04:24	
Diagnostics Check		OK	04:24	

Data Download Date/Time: 09/24/2016 20:48

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: WILLIAMS,SEAN
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 08/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 04:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:38

SUBJECT'S LAST NAME: PIERRE

DATE/TIME OF TEST: 09/18/2016 05:09:37

Diagnostics Check		OK	05:03
Air Blank		0.000	05:03
Control Test		0.078	05:03
Air Blank		0.000	05:04
Subject Sample	# 1	0.203	05:05
Breath Volume		1.492	05:05
Air Blank		0.000	05:05
Air Blank		0.000	05:07
Subject Sample	# 2	0.208	05:08
Breath Volume		1.425	05:08
Air Blank		0.000	05:08
Control Test		0.079	05:09
Air Blank		0.000	05:09
Diagnostics Check		OK	05:09

Data Download Date/Time: 09/24/2016 20:48

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: WASHINGTON,YVETTE
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 08/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: FERNANDEZLEAL

DATE/TIME OF TEST: 09/22/2016 03:07:34

Diagnostics Check		OK	03:01
Air Blank		0.000	03:01
Control Test		0.080	03:02
Air Blank		0.000	03:02
Subject Sample	# 1	0.172	03:03
Breath Volume		1.175	03:03
Air Blank		0.000	03:03
Air Blank		0.000	03:05
Subject Sample	# 2	0.171	03:06
Breath Volume		1.164	03:06
Air Blank		0.000	03:06
Control Test		0.079	03:06
Air Blank		0.000	03:07
Diagnostics Check		OK	03:07

Data Download Date/Time: 09/24/2016 20:48

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARNANDAN,INDROWTIE,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: MARTIN,JENNIFER
ARRESTING OFFICER AGENCY: Pembroke Pines PD
LAST AGENCY INSPECTION: 08/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 02:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:38

SUBJECT'S LAST NAME: CUMMINGS

DATE/TIME OF TEST: 09/23/2016 04:05:07

Diagnostics Check		OK	03:58
Air Blank		0.000	03:59
Control Test		0.079	03:59
Air Blank		0.000	03:59
Subject Sample	# 1	0.000	04:00
Breath Volume		2.140	04:00
Air Blank		0.000	04:00
Air Blank		0.000	04:02
Subject Sample	# 2	0.000	04:03
Breath Volume		2.269	04:03
Air Blank		0.000	04:04
Control Test		0.079	04:04
Air Blank		0.000	04:05
Diagnostics Check		OK	04:05

Data Download Date/Time: 09/24/2016 20:48

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARNANDAN,INDROWTIE,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: DIAZ,DANNY
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 08/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 00:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:38

SUBJECT'S LAST NAME: CACERES

DATE/TIME OF TEST: 09/24/2016 01:12:07

Diagnostics Check		OK	01:00	
Air Blank		0.000	01:00	
Control Test		0.080	01:00	
Air Blank		0.000	01:01	
Subject Sample	# 1	0.000	01:04	Volume Not Met
Breath Volume		0.351	01:04	
Air Blank		0.000	01:05	
Air Blank		0.000	01:07	
Subject Sample	# 2	0.191	01:07	
Breath Volume		2.695	01:07	
Air Blank		0.000	01:08	
Air Blank		0.000	01:10	
Subject Sample	# 3	0.196	01:10	
Breath Volume		2.644	01:10	
Air Blank		0.000	01:11	
Control Test		0.079	01:11	
Air Blank		0.000	01:12	
Diagnostics Check		OK	01:12	

Data Download Date/Time: 09/24/2016 20:48

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SERPHOS,STEVEN
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 08/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 01:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: THIBAULT

DATE/TIME OF TEST: 09/24/2016 02:28:41

Diagnostics Check		OK	02:20
Air Blank		0.000	02:21
Control Test		0.079	02:21
Air Blank		0.000	02:22
Subject Sample	# 1	0.172	02:22
Breath Volume		1.660	02:22
Air Blank		0.000	02:23
Air Blank		0.000	02:25
Subject Sample	# 2	0.165	02:27
Breath Volume		1.777	02:27
Air Blank		0.000	02:27
Control Test		0.079	02:28
Air Blank		0.000	02:28
Diagnostics Check		OK	02:28

Data Download Date/Time: 09/24/2016 20:48

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARNANDAN,INDROWTIE,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: MASSEY,AUSTIN
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 08/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 02:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: YOUNGBLOOD

DATE/TIME OF TEST: 09/24/2016 03:11:29

Diagnostics Check		OK	03:05
Air Blank		0.000	03:05
Control Test		0.079	03:05
Air Blank		0.000	03:06
Subject Sample	# 1	0.121	03:07
Breath Volume		1.332	03:07
Air Blank		0.000	03:07
Air Blank		0.000	03:09
Subject Sample	# 2	0.124	03:10
Breath Volume		2.414	03:10
Air Blank		0.000	03:10
Control Test		0.079	03:10
Air Blank		0.000	03:11
Diagnostics Check		OK	03:11

Data Download Date/Time: 09/24/2016 20:48

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARNANDAN,INDROWTIE,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FLINK,ANDREW
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 08/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 04:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:57

SUBJECT'S LAST NAME: ALLEN

DATE/TIME OF TEST: 09/24/2016 05:24:04

Diagnostics Check		OK	05:17
Air Blank		0.000	05:18
Control Test		0.080	05:18
Air Blank		0.000	05:19
Subject Sample	# 1	0.173	05:19
Breath Volume		1.269	05:19
Air Blank		0.000	05:20
Air Blank		0.000	05:22
Subject Sample	# 2	0.171	05:22
Breath Volume		2.070	05:22
Air Blank		0.000	05:23
Control Test		0.079	05:23
Air Blank		0.000	05:23
Diagnostics Check		OK	05:24

Data Download Date/Time: 09/24/2016 20:48

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DICIENZO,FRANCA,R
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: PARRISH,MATTHEW
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 08/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 14:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:45

SUBJECT'S LAST NAME: MEDINA

DATE/TIME OF TEST: 09/24/2016 15:45:08

Diagnostics Check		OK	15:38
Air Blank		0.000	15:39
Control Test		0.080	15:39
Air Blank		0.000	15:39
Subject Sample	# 1	0.265	15:40
Breath Volume		2.636	15:40
Air Blank		0.000	15:41
Air Blank		0.000	15:43
Subject Sample	# 2	0.264	15:43
Breath Volume		1.445	15:43
Air Blank		0.000	15:44
Control Test		0.079	15:44
Air Blank		0.000	15:45
Diagnostics Check		OK	15:45

Data Download Date/Time: 09/24/2016 20:48

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DICIENZO,FRANCA,R
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: HARITOS,PETE
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 08/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 17:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:55

SUBJECT'S LAST NAME: PATTERSON

DATE/TIME OF TEST: 09/24/2016 18:27:00

Diagnostics Check		OK	18:20
Air Blank		0.000	18:21
Control Test		0.079	18:21
Air Blank		0.000	18:22
Subject Sample	# 1	0.172	18:22
Breath Volume		2.843	18:22
Air Blank		0.000	18:23
Air Blank		0.000	18:25
Subject Sample	# 2	0.178	18:25
Breath Volume		2.566	18:25
Air Blank		0.000	18:26
Control Test		0.080	18:26
Air Blank		0.000	18:26
Diagnostics Check		OK	18:27

Data Download Date/Time: 09/24/2016 20:48

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 09/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:14

SUBJECT'S LAST NAME: KOON

DATE/TIME OF TEST: 09/23/2016 23:39:12

Diagnostics Check		OK	23:35	
Air Blank		0.000	23:35	
Control Test		0.079	23:35	
Air Blank		0.000	23:36	
Subject Sample	# 1	0.218	23:37	
Breath Volume		1.605	23:37	
Air Blank			23:37	RFI Detect
Air Blank			23:38	Purge Fail
Air Blank		0.000	23:39	

Data Download Date/Time: 10/20/2016 02:45

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 09/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:14

SUBJECT'S LAST NAME: KOON

DATE/TIME OF TEST: 09/23/2016 23:44:47

Diagnostics Check		OK	23:41	
Air Blank		0.000	23:41	
Control Test		0.077	23:42	
Air Blank		0.000	23:42	
Subject Sample	# 1	0.189	23:43	
Breath Volume		2.242	23:43	
Air Blank			23:43	Purge Fail
Air Blank			23:44	Purge Fail

Data Download Date/Time: 10/20/2016 02:45

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 09/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:14

SUBJECT'S LAST NAME: KOON

DATE/TIME OF TEST: 09/23/2016 23:48:37

Diagnostics Check	OK	23:47	
Air Blank		23:48	Ambient Fail
Air Blank	0.000	23:48	

Data Download Date/Time: 10/20/2016 02:45

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001073

INSTRUMENT REGISTERED TO: DADE CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AIKEN,DURRIEL,R
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: AIKEN,DURRIEL
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 08/17/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A3
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:28

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 09/16/2016 02:56:02

Diagnostics Check		OK	02:51	
Air Blank		0.000	02:51	
Control Test		0.078	02:52	
Air Blank		0.000	02:52	
Subject Sample	# 1		02:54	Subject Test Refused
Breath Volume		0.000	02:54	
Air Blank		0.000	02:55	
Control Test		0.079	02:55	
Air Blank		0.000	02:55	
Diagnostics Check		OK	02:56	

Data Download Date/Time: 09/19/2016 17:22

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001074

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIVERA,JOHN,L
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: RIVERA,JOHN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 08/19/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A1
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 19:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:45

SUBJECT'S LAST NAME: TURPEN

DATE/TIME OF TEST: 09/13/2016 20:30:24

Diagnostics Check		OK	20:23
Air Blank		0.000	20:24
Control Test		0.078	20:24
Air Blank		0.000	20:25
Subject Sample	# 1	0.000	20:26
Breath Volume		2.156	20:26
Air Blank		0.000	20:26
Air Blank		0.000	20:28
Subject Sample	# 2	0.000	20:29
Breath Volume		1.480	20:29
Air Blank		0.000	20:29
Control Test		0.077	20:29
Air Blank		0.000	20:30
Diagnostics Check		OK	20:30

Data Download Date/Time: 09/15/2016 10:24

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001075

INSTRUMENT REGISTERED TO: FHP TROOP E
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADDISON,RAYMOND,E
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: ADDISON,RAYMOND
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 08/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1824329
DRY GAS EXPIRY DATE: 03/09/2018
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: SEQUEIRA

DATE/TIME OF TEST: 09/02/2016 02:49:37

Diagnostics Check		OK	02:45	
Air Blank		0.000	02:46	
Control Test		0.079	02:46	
Air Blank		0.000	02:47	
Subject Sample	# 1		02:48	Subject Test Refused
Breath Volume		0.000	02:48	
Air Blank		0.000	02:48	
Control Test		0.078	02:49	
Air Blank		0.000	02:49	
Diagnostics Check		OK	02:49	

Data Download Date/Time: 09/16/2016 11:17

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001075

INSTRUMENT REGISTERED TO: FHP TROOP E
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADDISON,RAYMOND,E
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: SANCHEZ,MARGARITA
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 08/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1824329
DRY GAS EXPIRY DATE: 03/09/2018
VIOLATION CODE: DUI
ARREST TIME: 01:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: MARTINEZ RIVERA

DATE/TIME OF TEST: 09/04/2016 02:41:57

Diagnostics Check		OK	02:35
Air Blank		0.000	02:35
Control Test		0.077	02:36
Air Blank		0.000	02:36
Subject Sample	# 1	0.069	02:37
Breath Volume		1.867	02:37
Air Blank		0.000	02:38
Air Blank		0.000	02:40
Subject Sample	# 2	0.066	02:40
Breath Volume		2.273	02:40
Air Blank		0.000	02:41
Control Test		0.079	02:41
Air Blank		0.000	02:41
Diagnostics Check		OK	02:41

Data Download Date/Time: 09/16/2016 11:17

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001075

INSTRUMENT REGISTERED TO: FHP TROOP E
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADDISON,RAYMOND,E
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: MARRERO,JUAN
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 08/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1824329
DRY GAS EXPIRY DATE: 03/09/2018
VIOLATION CODE: DUI
ARREST TIME: 05:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:43

SUBJECT'S LAST NAME: MIJARES REMESAL

DATE/TIME OF TEST: 09/05/2016 06:10:03

Diagnostics Check		OK	06:03
Air Blank		0.000	06:04
Control Test		0.079	06:04
Air Blank		0.000	06:05
Subject Sample	# 1	0.165	06:05
Breath Volume		1.816	06:05
Air Blank		0.000	06:06
Air Blank		0.000	06:08
Subject Sample	# 2	0.162	06:08
Breath Volume		1.789	06:08
Air Blank		0.000	06:09
Control Test		0.078	06:09
Air Blank		0.000	06:09
Diagnostics Check		OK	06:10

Data Download Date/Time: 09/16/2016 11:17

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001075

INSTRUMENT REGISTERED TO: FHP TROOP E
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,ARGENIS,A
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: PEREZ,ARGENIS
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 08/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1824329
DRY GAS EXPIRY DATE: 03/09/2018
VIOLATION CODE: DUI
ARREST TIME: 08:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:40

SUBJECT'S LAST NAME: QUANT

DATE/TIME OF TEST: 09/05/2016 09:07:10

Diagnostics Check		OK	09:04	
Air Blank		0.000	09:04	
Control Test		0.079	09:05	
Air Blank		0.000	09:05	
Subject Sample	# 1		09:05	Subject Test Refused
Breath Volume		0.000	09:05	
Air Blank		0.000	09:06	
Control Test		0.079	09:06	
Air Blank		0.000	09:07	
Diagnostics Check		OK	09:07	

Data Download Date/Time: 09/16/2016 11:17

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001075

INSTRUMENT REGISTERED TO: FHP TROOP E
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRIEGO,CHARLENE,R
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: CABRERA,AMARANTO
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 08/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1824329
DRY GAS EXPIRY DATE: 03/09/2018
VIOLATION CODE: DUI
ARREST TIME: 23:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:01

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 09/14/2016 00:31:59

Diagnostics Check		OK	00:25
Air Blank		0.000	00:25
Control Test		0.078	00:25
Air Blank		0.000	00:26
Subject Sample	# 1	0.112	00:27
Breath Volume		1.523	00:27
Air Blank		0.000	00:28
Air Blank		0.000	00:30
Subject Sample	# 2	0.109	00:30
Breath Volume		1.539	00:30
Air Blank		0.000	00:31
Control Test		0.078	00:31
Air Blank		0.000	00:31
Diagnostics Check		OK	00:31

Data Download Date/Time: 09/16/2016 11:17

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001075

INSTRUMENT REGISTERED TO: FHP TROOP E
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ-CORTINA,RAUL,H
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: GONZALEZ-CORTINA,RAUL
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 08/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1824329
DRY GAS EXPIRY DATE: 03/09/2018
VIOLATION CODE: DUI
ARREST TIME: 03:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: CARILLO MANZANO

DATE/TIME OF TEST: 09/15/2016 04:24:27

Diagnostics Check		OK	04:20	
Air Blank		0.000	04:21	
Control Test		0.079	04:21	
Air Blank		0.000	04:22	
Subject Sample	# 1		04:23	Subject Test Refused
Breath Volume		0.000	04:23	
Air Blank		0.000	04:23	
Control Test		0.077	04:23	
Air Blank		0.000	04:24	
Diagnostics Check		OK	04:24	

Data Download Date/Time: 09/16/2016 11:17

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001075

INSTRUMENT REGISTERED TO: FHP TROOP E
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALLO,DENNIS,
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: GALLO,DENNIS
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 09/16/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1694353
DRY GAS EXPIRY DATE: 06/10/2017
VIOLATION CODE: DUI
ARREST TIME: 00:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:03

SUBJECT'S LAST NAME: GONZALEZ GONZALEZ

DATE/TIME OF TEST: 09/23/2016 01:45:38

Diagnostics Check		OK	01:37
Air Blank		0.000	01:37
Control Test		0.077	01:37
Air Blank		0.000	01:38
Subject Sample	# 1	0.217	01:41
Breath Volume		2.503	01:41
Air Blank		0.000	01:41
Air Blank		0.000	01:43
Subject Sample	# 2	0.226	01:44
Breath Volume		2.367	01:44
Air Blank		0.000	01:44
Control Test		0.078	01:45
Air Blank		0.000	01:45
Diagnostics Check		OK	01:45

Data Download Date/Time: 10/31/2016 08:40

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001075

INSTRUMENT REGISTERED TO: FHP TROOP E
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANTANA LOPEZ,ERNESTO,
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: SANTANA LOPEZ,ERNESTO
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 09/16/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1824329
DRY GAS EXPIRY DATE: 03/09/2018
VIOLATION CODE: DUI
ARREST TIME: 14:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:20

SUBJECT'S LAST NAME: FIGUEROA

DATE/TIME OF TEST: 09/24/2016 18:02:48

Diagnostics Check		OK	17:47	
Air Blank		0.000	17:47	
Control Test		0.078	17:47	
Air Blank		0.000	17:48	
Subject Sample	# 1	0.000	17:50	
Breath Volume		1.121	17:50	
Air Blank		0.000	17:50	
Air Blank		0.000	17:52	
Subject Sample	# 2	0.000	17:55	Volume Not Met
Breath Volume		1.011	17:55	
Air Blank		0.000	17:56	
Air Blank		0.000	17:58	
Subject Sample	# 3	0.000	18:01	Volume Not Met
Breath Volume		0.890	18:01	
Air Blank		0.000	18:01	
Control Test		0.077	18:02	
Air Blank		0.000	18:02	
Diagnostics Check		OK	18:02	

Data Download Date/Time: 10/31/2016 08:40

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001075

INSTRUMENT REGISTERED TO: FHP TROOP E
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIVERA,NEPHTALI,
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: RIVERA,NEPHTALI
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 09/16/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1824329
DRY GAS EXPIRY DATE: 03/09/2018
VIOLATION CODE: DUI
ARREST TIME: 05:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:05

SUBJECT'S LAST NAME: HERNANDEZ ALONSO

DATE/TIME OF TEST: 09/25/2016 06:37:12

Diagnostics Check		OK	06:30
Air Blank		0.000	06:31
Control Test		0.077	06:31
Air Blank		0.000	06:31
Subject Sample	# 1	0.201	06:32
Breath Volume		2.312	06:32
Air Blank		0.000	06:33
Air Blank		0.000	06:35
Subject Sample	# 2	0.190	06:35
Breath Volume		1.550	06:35
Air Blank		0.000	06:36
Control Test		0.078	06:36
Air Blank		0.000	06:37
Diagnostics Check		OK	06:37

Data Download Date/Time: 10/31/2016 08:40

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001075

INSTRUMENT REGISTERED TO: FHP TROOP E
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO,CARLOS,J
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: ROSARIO,CARLOS
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 09/16/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1824329
DRY GAS EXPIRY DATE: 03/09/2018
VIOLATION CODE: DUI
ARREST TIME: 02:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: MURILLO

DATE/TIME OF TEST: 09/30/2016 03:43:15

Diagnostics Check		OK	03:30
Air Blank		0.000	03:31
Control Test		0.079	03:31
Air Blank		0.000	03:32
Subject Sample	# 1	0.196	03:33
Breath Volume		1.187	03:33
Air Blank		0.000	03:33
Air Blank		0.000	03:35
Subject Sample	# 2	0.243	03:36
Breath Volume		1.714	03:36
Air Blank		0.000	03:37
Air Blank		0.000	03:38
Subject Sample	# 3	0.237	03:41
Breath Volume		1.867	03:41
Air Blank		0.000	03:42
Control Test		0.078	03:42
Air Blank		0.000	03:43
Diagnostics Check		OK	03:43

Data Download Date/Time: 10/31/2016 08:40

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001077

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONEY,MARY,L
BREATH TEST OPERATOR AGENCY: Hardee County SO
ARRESTING OFFICER: BRANDEBERRY,JEREMY
ARRESTING OFFICER AGENCY: Hardee County SO
LAST AGENCY INSPECTION: 08/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 01:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 09/19/2016 01:50:54

Diagnostics Check		OK	01:46	
Air Blank		0.000	01:47	
Control Test		0.083	01:47	
Air Blank		0.000	01:48	
Subject Sample	# 1		01:49	Subject Test Refused
Breath Volume		0.000	01:49	
Air Blank		0.000	01:49	
Control Test		0.082	01:50	
Air Blank		0.000	01:50	
Diagnostics Check		OK	01:50	

Data Download Date/Time: 09/28/2016 23:48

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001077

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARAUJO,NESTOR,J
BREATH TEST OPERATOR AGENCY: Hardee County SO
ARRESTING OFFICER: WARD,SHANE
ARRESTING OFFICER AGENCY: Hardee County SO
LAST AGENCY INSPECTION: 08/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 20:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:40

SUBJECT'S LAST NAME: TREMBLAY

DATE/TIME OF TEST: 09/20/2016 21:09:32

Diagnostics Check		OK	21:01
Air Blank		0.000	21:02
Control Test		0.082	21:02
Air Blank		0.000	21:03
Subject Sample	# 1	0.165	21:04
Breath Volume		1.390	21:04
Air Blank		0.000	21:05
Air Blank		0.000	21:07
Subject Sample	# 2	0.166	21:07
Breath Volume		1.925	21:07
Air Blank		0.000	21:08
Control Test		0.081	21:08
Air Blank		0.000	21:09
Diagnostics Check		OK	21:09

Data Download Date/Time: 09/28/2016 23:48

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRUBAUGH,ANDRE,J
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ZOLLOTA,VLADIMIR
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 08/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 01:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: CARABALLO

DATE/TIME OF TEST: 09/01/2016 02:44:27

Diagnostics Check	OK	02:43	
Air Blank		02:43	RFI Detect
Air Blank	0.000	02:44	

Data Download Date/Time: 09/14/2016 09:31

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRUBAUGH,ANDRE,J
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ZOLLOTA,VLADIMIR
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 08/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 01:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: CARABALLO

DATE/TIME OF TEST: 09/01/2016 02:49:29

Diagnostics Check		OK	02:46	
Air Blank		0.000	02:47	
Control Test		0.079	02:47	
Air Blank		0.000	02:47	
Subject Sample	# 1		02:48	Subject Test Refused
Breath Volume		0.000	02:48	
Air Blank		0.000	02:48	
Control Test		0.080	02:48	
Air Blank		0.000	02:49	
Diagnostics Check		OK	02:49	

Data Download Date/Time: 09/14/2016 09:31

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COX,LENARD,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ALLI,ANTHONY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 08/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 22:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:56

SUBJECT'S LAST NAME: RUSSO

DATE/TIME OF TEST: 09/01/2016 23:25:32

Diagnostics Check		OK	23:19
Air Blank		0.000	23:20
Control Test		0.081	23:20
Air Blank		0.000	23:20
Subject Sample	# 1	0.000	23:21
Breath Volume		2.593	23:21
Air Blank		0.000	23:21
Air Blank		0.000	23:23
Subject Sample	# 2	0.000	23:24
Breath Volume		1.812	23:24
Air Blank		0.000	23:24
Control Test		0.081	23:24
Air Blank		0.000	23:25
Diagnostics Check		OK	23:25

Data Download Date/Time: 09/14/2016 09:31

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZOLLOTA,VLADIMIR,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: COX,LENARD
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 08/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 00:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: GRANT

DATE/TIME OF TEST: 09/02/2016 00:48:26

Diagnostics Check		OK	00:41
Air Blank		0.000	00:41
Control Test		0.080	00:42
Air Blank		0.000	00:42
Subject Sample	# 1	0.179	00:43
Breath Volume		1.394	00:43
Air Blank		0.000	00:44
Air Blank		0.000	00:46
Subject Sample	# 2	0.175	00:46
Breath Volume		1.375	00:46
Air Blank		0.000	00:47
Control Test		0.078	00:47
Air Blank		0.000	00:48
Diagnostics Check		OK	00:48

Data Download Date/Time: 09/14/2016 09:31

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COX,LENARD,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: GRUBAUGH,ANDRE
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 08/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 21:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:45

SUBJECT'S LAST NAME: KNIGHT

DATE/TIME OF TEST: 09/03/2016 21:25:22

Diagnostics Check		OK	21:17
Air Blank		0.000	21:17
Control Test		0.079	21:18
Air Blank		0.000	21:18
Subject Sample	# 1	0.150	21:19
Breath Volume		1.894	21:19
Air Blank		0.000	21:19
Air Blank		0.000	21:21
Subject Sample	# 2	0.161	21:23
Breath Volume		2.378	21:23
Air Blank		0.000	21:24
Control Test		0.079	21:24
Air Blank		0.000	21:25
Diagnostics Check		OK	21:25

Data Download Date/Time: 09/14/2016 09:31

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COX,LENARD,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ALLI,ANTHONY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 08/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 23:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:07

SUBJECT'S LAST NAME: FISHER

DATE/TIME OF TEST: 09/04/2016 00:35:57

Diagnostics Check		OK	00:27
Air Blank		0.000	00:28
Control Test		0.079	00:28
Air Blank		0.000	00:28
Subject Sample	# 1	0.239	00:29
Breath Volume		1.851	00:29
Air Blank		0.000	00:30
Air Blank		0.000	00:31
Subject Sample	# 2	0.256	00:34
Breath Volume		2.878	00:34
Air Blank		0.000	00:35
Control Test		0.080	00:35
Air Blank		0.000	00:35
Diagnostics Check		OK	00:35

Data Download Date/Time: 09/14/2016 09:31

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ACRI,MICHEAL,N
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: SMITH,EMILY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 08/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 05:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:45

SUBJECT'S LAST NAME: NOBLES

DATE/TIME OF TEST: 09/04/2016 06:14:14

Diagnostics Check		OK	06:05
Air Blank		0.000	06:05
Control Test		0.080	06:06
Air Blank		0.000	06:06
Subject Sample	# 1	0.096	06:09
Breath Volume		1.695	06:09
Air Blank		0.000	06:09
Air Blank		0.000	06:11
Subject Sample	# 2	0.091	06:12
Breath Volume		2.539	06:12
Air Blank		0.000	06:13
Control Test		0.079	06:13
Air Blank		0.000	06:14
Diagnostics Check		OK	06:14

Data Download Date/Time: 09/14/2016 09:31

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLARK,BRENT,D
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CLARK,BRENT
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 08/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 16:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:15

SUBJECT'S LAST NAME: COVINGTON

DATE/TIME OF TEST: 09/06/2016 17:49:40

Diagnostics Check		OK	17:35	
Air Blank		0.000	17:35	
Control Test		0.081	17:36	
Air Blank		0.000	17:36	
Subject Sample	# 1	0.227	17:39	
Breath Volume		1.148	17:39	
Air Blank		0.000	17:39	
Air Blank		0.000	17:41	
Subject Sample	# 2	0.204	17:44	Volume Not Met
Breath Volume		0.566	17:44	
Air Blank		0.000	17:45	
Air Blank		0.000	17:47	
Subject Sample	# 3	0.236	17:48	
Breath Volume		1.839	17:48	
Air Blank		0.000	17:48	
Control Test		0.080	17:49	
Air Blank		0.000	17:49	
Diagnostics Check		OK	17:49	

Data Download Date/Time: 09/14/2016 09:31

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KARAYIANES,MICHAEL,F
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ZOLLOTA,VLADIMIR
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 08/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 21:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:22

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 09/08/2016 22:50:10

Diagnostics Check		OK	22:42
Air Blank		0.000	22:43
Control Test		0.080	22:43
Air Blank		0.000	22:43
Subject Sample	# 1	0.215	22:44
Breath Volume		3.867	22:44
Air Blank		0.000	22:45
Air Blank		0.000	22:47
Subject Sample	# 2	0.195	22:48
Breath Volume		2.121	22:48
Air Blank		0.000	22:49
Control Test		0.079	22:49
Air Blank		0.000	22:50
Diagnostics Check		OK	22:50

Data Download Date/Time: 09/14/2016 09:31

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KARAYIANES,MICHAEL,F
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: COX,LENARD
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 08/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 01:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:07

SUBJECT'S LAST NAME: VANVLEET

DATE/TIME OF TEST: 09/09/2016 02:35:48

Diagnostics Check		OK	02:27
Air Blank		0.000	02:27
Control Test		0.080	02:28
Air Blank		0.000	02:28
Subject Sample	# 1	0.193	02:31
Breath Volume		2.476	02:31
Air Blank		0.000	02:31
Air Blank		0.000	02:33
Subject Sample	# 2	0.194	02:34
Breath Volume		2.011	02:34
Air Blank		0.000	02:34
Control Test		0.078	02:35
Air Blank		0.000	02:35
Diagnostics Check		OK	02:35

Data Download Date/Time: 09/14/2016 09:31

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KARAYIANES,MICHAEL,F
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ALLI,ANTHONY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 08/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 04:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:14

SUBJECT'S LAST NAME: FAIRLEY

DATE/TIME OF TEST: 09/09/2016 04:41:12

Diagnostics Check		OK	04:34
Air Blank		0.000	04:35
Control Test		0.080	04:35
Air Blank		0.000	04:36
Subject Sample	# 1	0.013	04:36
Breath Volume		1.601	04:36
Air Blank		0.000	04:37
Air Blank		0.000	04:39
Subject Sample	# 2	0.011	04:39
Breath Volume		1.394	04:39
Air Blank		0.000	04:40
Control Test		0.080	04:40
Air Blank		0.000	04:41
Diagnostics Check		OK	04:41

Data Download Date/Time: 09/14/2016 09:31

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KARAYIANES,MICHAEL,F
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: GRUBAUGH,ANDRE
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 08/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 18:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:11

SUBJECT'S LAST NAME: DAVIDSON

DATE/TIME OF TEST: 09/09/2016 19:38:39

Diagnostics Check		OK	19:31
Air Blank		0.000	19:32
Control Test		0.079	19:32
Air Blank		0.000	19:33
Subject Sample	# 1	0.194	19:34
Breath Volume		1.445	19:34
Air Blank		0.000	19:34
Air Blank		0.000	19:36
Subject Sample	# 2	0.201	19:37
Breath Volume		1.320	19:37
Air Blank		0.000	19:37
Control Test		0.078	19:38
Air Blank		0.000	19:38
Diagnostics Check		OK	19:38

Data Download Date/Time: 09/14/2016 09:31

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KARAYIANES,MICHAEL,F
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ZOLLOTA,VLADIMIR
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 08/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:32

SUBJECT'S LAST NAME: BELEN

DATE/TIME OF TEST: 09/10/2016 03:01:21

Diagnostics Check		OK	02:52	
Air Blank		0.000	02:53	
Control Test		0.080	02:53	
Air Blank		0.000	02:54	
Subject Sample	# 1		02:57	No Sample Provided
Breath Volume		0.000	02:57	
Air Blank		0.000	02:57	
Air Blank		0.000	02:59	
Subject Sample	# 2		02:59	Subject Test Refused
Breath Volume		0.000	02:59	
Air Blank		0.000	03:00	
Control Test		0.080	03:00	
Air Blank		0.000	03:01	
Diagnostics Check		OK	03:01	

Data Download Date/Time: 09/14/2016 09:31

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRUBAUGH,ANDRE,J
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: KARAYIANES,M
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 08/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 03:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: RUSSELL

DATE/TIME OF TEST: 09/11/2016 04:04:50

Diagnostics Check		OK	04:01	
Air Blank		0.000	04:02	
Control Test		0.079	04:02	
Air Blank		0.000	04:02	
Subject Sample	# 1		04:03	Subject Test Refused
Breath Volume		0.000	04:03	
Air Blank		0.000	04:03	
Control Test		0.079	04:04	
Air Blank		0.000	04:04	
Diagnostics Check		OK	04:04	

Data Download Date/Time: 09/14/2016 09:31

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALKER,CALVIN,D
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: MORGAN,QUENTINLEE
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 08/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 20:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:35

SUBJECT'S LAST NAME: UOGINTIENE

DATE/TIME OF TEST: 09/12/2016 21:13:29

Diagnostics Check		OK	21:07
Air Blank		0.000	21:07
Control Test		0.080	21:08
Air Blank		0.000	21:08
Subject Sample	# 1	0.221	21:09
Breath Volume		2.246	21:09
Air Blank		0.000	21:09
Air Blank		0.000	21:11
Subject Sample	# 2	0.230	21:11
Breath Volume		2.312	21:11
Air Blank		0.000	21:12
Control Test		0.079	21:12
Air Blank		0.000	21:13
Diagnostics Check		OK	21:13

Data Download Date/Time: 09/14/2016 09:31

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTER,MATTHEW,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CARTER,MATTHEW
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 08/30/2016

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 00:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:01

SUBJECT'S LAST NAME: DUTCHER

DATE/TIME OF TEST: 09/14/2016 01:38:15

Diagnostics Check		OK	01:34	
Air Blank		0.000	01:34	
Control Test		0.080	01:35	
Air Blank		0.000	01:35	
Subject Sample	# 1		01:36	Subject Test Refused
Breath Volume		0.000	01:36	
Air Blank		0.000	01:37	
Control Test		0.080	01:37	
Air Blank		0.000	01:38	
Diagnostics Check		OK	01:38	

Data Download Date/Time: 09/14/2016 09:31

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZOLLOTA,VLADIMIR,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: COX,LENARD
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 09/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 23:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: HILLSMAN

DATE/TIME OF TEST: 09/16/2016 00:16:18

Diagnostics Check		OK	00:10	
Air Blank		0.000	00:11	
Control Test		0.080	00:11	
Air Blank		0.000	00:12	
Subject Sample	# 1		00:14	Subject Test Refused
Breath Volume		0.000	00:14	
Air Blank		0.000	00:15	
Control Test		0.079	00:15	
Air Blank		0.000	00:16	
Diagnostics Check		OK	00:16	

Data Download Date/Time: 10/12/2016 10:37

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COX,LENARD,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: KARAYIANES,MICHAEL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 09/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 02:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:24

SUBJECT'S LAST NAME: FLESNER

DATE/TIME OF TEST: 09/17/2016 02:56:41

Diagnostics Check		OK	02:45
Air Blank		0.000	02:45
Control Test		0.081	02:45
Air Blank		0.000	02:46
Subject Sample	# 1	0.223	02:46
Breath Volume		1.851	02:46
Air Blank		0.000	02:47
Air Blank		0.000	02:49
Subject Sample	# 2	0.198	02:52
Breath Volume		1.929	02:52
Air Blank		0.000	02:52
Air Blank		0.000	02:54
Subject Sample	# 3	0.219	02:55
Breath Volume		1.902	02:55
Air Blank		0.000	02:55
Control Test		0.079	02:56
Air Blank		0.000	02:56
Diagnostics Check		OK	02:56

Data Download Date/Time: 10/12/2016 10:37

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KARAYIANES,MICHAEL,F
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: GRUBAUGH,ANDRE
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 09/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 21:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:17

SUBJECT'S LAST NAME: BARRETT

DATE/TIME OF TEST: 09/17/2016 22:42:22

Diagnostics Check		OK	22:37	
Air Blank		0.000	22:37	
Control Test		0.079	22:38	
Air Blank		0.000	22:38	
Subject Sample	# 1		22:41	Subject Test Refused
Breath Volume		0.000	22:41	
Air Blank		0.000	22:41	
Control Test		0.078	22:41	
Air Blank		0.000	22:42	
Diagnostics Check		OK	22:42	

Data Download Date/Time: 10/12/2016 10:37

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCE JR,SCOTT,A
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: PIERCE JR,SCOTT
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 09/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 13:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:10

SUBJECT'S LAST NAME: PALMER

DATE/TIME OF TEST: 09/20/2016 14:38:52

Diagnostics Check		OK	14:30
Air Blank		0.000	14:31
Control Test		0.079	14:31
Air Blank		0.000	14:32
Subject Sample	# 1	0.175	14:34
Breath Volume		1.136	14:34
Air Blank		0.000	14:34
Air Blank		0.000	14:36
Subject Sample	# 2	0.178	14:37
Breath Volume		2.246	14:37
Air Blank		0.000	14:37
Control Test		0.079	14:38
Air Blank		0.000	14:38
Diagnostics Check		OK	14:38

Data Download Date/Time: 10/12/2016 10:37

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COX,LENARD,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ZOLLOTA,VLADIMIR
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 09/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 01:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 09/22/2016 01:53:38

Diagnostics Check		OK	01:50	
Air Blank		0.000	01:51	
Control Test		0.079	01:51	
Air Blank		0.000	01:52	
Subject Sample	# 1		01:52	Subject Test Refused
Breath Volume		0.000	01:52	
Air Blank		0.000	01:52	
Control Test		0.080	01:53	
Air Blank		0.000	01:53	
Diagnostics Check		OK	01:53	

Data Download Date/Time: 10/12/2016 10:37

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KARAYIANES,MICHAEL,F
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ALLI,ANTHONY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 09/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 03:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:02

SUBJECT'S LAST NAME: GREEN

DATE/TIME OF TEST: 09/22/2016 04:30:22

Diagnostics Check		OK	04:23
Air Blank		0.000	04:23
Control Test		0.080	04:24
Air Blank		0.000	04:24
Subject Sample	# 1	0.311	04:25
Breath Volume		2.707	04:25
Air Blank		0.000	04:25
Air Blank		0.000	04:27
Subject Sample	# 2	0.309	04:28
Breath Volume		2.187	04:28
Air Blank		0.000	04:29
Control Test		0.077	04:29
Air Blank		0.000	04:30
Diagnostics Check		OK	04:30

Data Download Date/Time: 10/12/2016 10:37

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COX,LENARD,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: COX,LENARD
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 09/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 22:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 09/22/2016 23:12:09

Diagnostics Check		OK	23:06
Air Blank		0.000	23:06
Control Test		0.080	23:07
Air Blank		0.000	23:07
Subject Sample	# 1	0.115	23:07
Breath Volume		1.562	23:07
Air Blank		0.000	23:08
Air Blank		0.000	23:10
Subject Sample	# 2	0.112	23:10
Breath Volume		2.128	23:10
Air Blank		0.000	23:11
Control Test		0.080	23:11
Air Blank		0.000	23:12
Diagnostics Check		OK	23:12

Data Download Date/Time: 10/12/2016 10:37

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRUBAUGH,ANDRE,J
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: KARAYIANES,MICHAEL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 09/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 22:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: FOGERTY

DATE/TIME OF TEST: 09/23/2016 23:44:27

Diagnostics Check		OK	23:38
Air Blank		0.000	23:38
Control Test		0.080	23:39
Air Blank		0.000	23:39
Subject Sample	# 1	0.126	23:40
Breath Volume		2.109	23:40
Air Blank		0.000	23:40
Air Blank		0.000	23:42
Subject Sample	# 2	0.123	23:42
Breath Volume		3.046	23:42
Air Blank		0.000	23:43
Control Test		0.078	23:43
Air Blank		0.000	23:44
Diagnostics Check		OK	23:44

Data Download Date/Time: 10/12/2016 10:37

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KARAYIANES,MICHAEL,F
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: GRUBAUGH,ANDRE
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 09/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 03:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:52

SUBJECT'S LAST NAME: BRINKLEY

DATE/TIME OF TEST: 09/24/2016 04:20:56

Diagnostics Check		OK	04:12	
Air Blank		0.000	04:12	
Control Test		0.080	04:13	
Air Blank		0.000	04:13	
Subject Sample	# 1	0.171	04:16	Volume Not Met
Breath Volume		0.066	04:16	
Air Blank		0.000	04:17	
Air Blank		0.000	04:19	
Subject Sample	# 2		04:19	RFI Detect
Breath Volume		0.000	04:19	
Air Blank		0.000	04:20	
Control Test		0.078	04:20	
Air Blank		0.000	04:20	
Diagnostics Check		OK	04:20	

Data Download Date/Time: 10/12/2016 10:37

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KARAYIANES,MICHAEL,F
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: GRUBAUGH,ANDRE
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 09/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 03:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:52

SUBJECT'S LAST NAME: BRINKLEY

DATE/TIME OF TEST: 09/24/2016 04:33:57

Diagnostics Check		OK	04:22	
Air Blank		0.000	04:22	
Control Test		0.079	04:23	
Air Blank		0.000	04:23	
Subject Sample	# 1	0.170	04:26	Volume Not Met
Breath Volume		0.035	04:26	
Air Blank		0.000	04:27	
Air Blank		0.000	04:29	
Subject Sample	# 2	0.160	04:32	Volume Not Met
Breath Volume		0.632	04:32	
Air Blank		0.000	04:33	
Control Test		0.076	04:33	
Air Blank		0.000	04:33	
Diagnostics Check		OK	04:33	

Data Download Date/Time: 10/12/2016 10:37

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COX,LENARD,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ALLI,ANTHONY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 09/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 18:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:27

SUBJECT'S LAST NAME: SCHNEIDER

DATE/TIME OF TEST: 09/24/2016 18:57:11

Diagnostics Check		OK	18:49
Air Blank		0.000	18:49
Control Test		0.080	18:50
Air Blank		0.000	18:50
Subject Sample	# 1	0.000	18:51
Breath Volume		3.640	18:51
Air Blank		0.000	18:52
Air Blank		0.000	18:54
Subject Sample	# 2	0.000	18:55
Breath Volume		3.769	18:55
Air Blank		0.000	18:56
Control Test		0.079	18:56
Air Blank		0.000	18:57
Diagnostics Check		OK	18:57

Data Download Date/Time: 10/12/2016 10:37

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COX,LENARD,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ZOLLOTA,VLADIMIR
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 09/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 18:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:10

SUBJECT'S LAST NAME: SARBER

DATE/TIME OF TEST: 09/24/2016 19:32:47

Diagnostics Check	OK	19:32	
Air Blank		19:32	RFI Detect
Air Blank	0.000	19:32	

Data Download Date/Time: 10/12/2016 10:37

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COX,LENARD,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ZOLLOTA,VLADIMIR
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 09/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 18:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:10

SUBJECT'S LAST NAME: SARBER

DATE/TIME OF TEST: 09/24/2016 19:40:11

Diagnostics Check		OK	19:33
Air Blank		0.000	19:34
Control Test		0.079	19:34
Air Blank		0.000	19:35
Subject Sample	# 1	0.000	19:35
Breath Volume		1.253	19:35
Air Blank		0.000	19:36
Air Blank		0.000	19:38
Subject Sample	# 2	0.000	19:38
Breath Volume		2.011	19:38
Air Blank		0.000	19:39
Control Test		0.079	19:39
Air Blank		0.000	19:40
Diagnostics Check		OK	19:40

Data Download Date/Time: 10/12/2016 10:37

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COX,LENARD,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: COX,LENARD
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 09/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 00:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: LARAYEBRA

DATE/TIME OF TEST: 09/25/2016 01:05:56

Diagnostics Check		OK	00:58
Air Blank		0.000	00:59
Control Test		0.080	00:59
Air Blank		0.000	01:00
Subject Sample	# 1	0.104	01:01
Breath Volume		1.812	01:01
Air Blank		0.000	01:02
Air Blank		0.000	01:04
Subject Sample	# 2	0.100	01:04
Breath Volume		1.585	01:04
Air Blank		0.000	01:05
Control Test		0.078	01:05
Air Blank		0.000	01:05
Diagnostics Check		OK	01:05

Data Download Date/Time: 10/12/2016 10:37

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KARAYIANES,MICHAEL,F
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: COX,LENARD
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 09/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 01:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: GRANDE

DATE/TIME OF TEST: 09/29/2016 01:53:34

Diagnostics Check		OK	01:51	
Air Blank		0.000	01:51	
Control Test		0.080	01:51	
Air Blank		0.000	01:52	
Subject Sample	# 1		01:53	Sequence Aborted
Breath Volume		0.000	01:53	
Air Blank		0.000	01:53	

Data Download Date/Time: 10/12/2016 10:37

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KARAYIANES,MICHAEL,F
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: COX,LENARD
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 09/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 01:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:54

SUBJECT'S LAST NAME: GRANDE

DATE/TIME OF TEST: 09/29/2016 02:21:25

Diagnostics Check		OK	02:14
Air Blank		0.000	02:15
Control Test		0.079	02:15
Air Blank		0.000	02:16
Subject Sample	# 1	0.160	02:16
Breath Volume		3.246	02:16
Air Blank		0.000	02:17
Air Blank		0.000	02:19
Subject Sample	# 2	0.156	02:19
Breath Volume		1.636	02:19
Air Blank		0.000	02:20
Control Test		0.080	02:20
Air Blank		0.000	02:21
Diagnostics Check		OK	02:21

Data Download Date/Time: 10/12/2016 10:38

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZOLLOTA,VLADIMIR,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: KARAYIANES,MICHAEL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 09/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 02:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: SHALER

DATE/TIME OF TEST: 09/30/2016 02:47:30

Diagnostics Check		OK	02:41
Air Blank		0.000	02:41
Control Test		0.078	02:42
Air Blank		0.000	02:42
Subject Sample	# 1	0.189	02:43
Breath Volume		2.750	02:43
Air Blank		0.000	02:43
Air Blank		0.000	02:45
Subject Sample	# 2	0.186	02:45
Breath Volume		2.542	02:45
Air Blank		0.000	02:46
Control Test		0.079	02:46
Air Blank		0.000	02:47
Diagnostics Check		OK	02:47

Data Download Date/Time: 10/12/2016 10:38

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BALDIE,BRYAN,A
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: GOSIER,M
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 08/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A3
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:52

SUBJECT'S LAST NAME: PALERMO

DATE/TIME OF TEST: 09/04/2016 00:19:52

Diagnostics Check		OK	00:12
Air Blank		0.000	00:13
Control Test		0.078	00:13
Air Blank		0.000	00:14
Subject Sample	# 1	0.135	00:14
Breath Volume		1.324	00:14
Air Blank		0.000	00:15
Air Blank		0.000	00:17
Subject Sample	# 2	0.139	00:18
Breath Volume		1.503	00:18
Air Blank		0.000	00:18
Control Test		0.079	00:19
Air Blank		0.000	00:19
Diagnostics Check		OK	00:19

Data Download Date/Time: 09/19/2016 06:25

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FREUDENTHAL,RYAN,T
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: GOSIER,MATT
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 08/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A3
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: FILIPI

DATE/TIME OF TEST: 09/04/2016 04:18:08

Diagnostics Check		OK	04:10
Air Blank		0.000	04:10
Control Test		0.078	04:11
Air Blank		0.000	04:11
Subject Sample	# 1	0.130	04:13
Breath Volume		2.125	04:13
Air Blank		0.000	04:14
Air Blank		0.000	04:15
Subject Sample	# 2	0.113	04:16
Breath Volume		1.507	04:16
Air Blank		0.000	04:17
Control Test		0.078	04:17
Air Blank		0.000	04:18
Diagnostics Check		OK	04:18

Data Download Date/Time: 09/19/2016 06:25

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLICKENSDORF,MICHAEL,R
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: GOSIER,MATTHEW
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 08/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A3
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:35

SUBJECT'S LAST NAME: DELGRANGE

DATE/TIME OF TEST: 09/04/2016 06:04:08

Diagnostics Check		OK	05:56
Air Blank		0.000	05:57
Control Test		0.079	05:57
Air Blank		0.000	05:57
Subject Sample	# 1	0.078	05:58
Breath Volume		1.917	05:58
Air Blank		0.000	05:59
Air Blank		0.000	06:01
Subject Sample	# 2	0.077	06:02
Breath Volume		1.500	06:02
Air Blank		0.000	06:03
Control Test		0.080	06:03
Air Blank		0.000	06:04
Diagnostics Check		OK	06:04

Data Download Date/Time: 09/19/2016 06:25

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FREUDENTHAL,RYAN,T
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: GOSIER,MATT
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 08/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A3
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:27

SUBJECT'S LAST NAME: WYNDHAM

DATE/TIME OF TEST: 09/05/2016 03:54:25

Diagnostics Check		OK	03:48
Air Blank		0.000	03:48
Control Test		0.078	03:48
Air Blank		0.000	03:49
Subject Sample	# 1	0.187	03:49
Breath Volume		3.597	03:49
Air Blank		0.000	03:50
Air Blank		0.000	03:52
Subject Sample	# 2	0.191	03:52
Breath Volume		3.085	03:52
Air Blank		0.000	03:53
Control Test		0.077	03:53
Air Blank		0.000	03:54
Diagnostics Check		OK	03:54

Data Download Date/Time: 09/19/2016 06:25

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 08/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A3
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:58

SUBJECT'S LAST NAME: SHEEHAN

DATE/TIME OF TEST: 09/08/2016 22:25:50

Diagnostics Check		OK	22:19
Air Blank		0.000	22:19
Control Test		0.078	22:20
Air Blank		0.000	22:20
Subject Sample	# 1	0.233	22:21
Breath Volume		2.480	22:21
Air Blank		0.000	22:22
Air Blank		0.000	22:23
Subject Sample	# 2	0.251	22:24
Breath Volume		2.535	22:24
Air Blank		0.000	22:24
Control Test		0.078	22:25
Air Blank		0.000	22:25
Diagnostics Check		OK	22:25

Data Download Date/Time: 09/19/2016 06:25

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: GOSIER,MATTHEW
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 08/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A3
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:14

SUBJECT'S LAST NAME: ULMER

DATE/TIME OF TEST: 09/10/2016 03:39:14

Diagnostics Check		OK	03:35	
Air Blank		0.000	03:35	
Control Test		0.078	03:35	
Air Blank		0.000	03:36	
Subject Sample	# 1		03:37	Subject Test Refused
Breath Volume		0.000	03:37	
Air Blank		0.000	03:38	
Control Test		0.078	03:38	
Air Blank		0.000	03:39	
Diagnostics Check		OK	03:39	

Data Download Date/Time: 09/19/2016 06:25

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FREUDENTHAL,RYAN,T
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: GOSIER,MATT
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 08/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A3
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 05:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:49

SUBJECT'S LAST NAME: BELLOSI

DATE/TIME OF TEST: 09/10/2016 06:16:41

Diagnostics Check		OK	06:09
Air Blank		0.000	06:10
Control Test		0.078	06:10
Air Blank		0.000	06:11
Subject Sample	# 1	0.000	06:12
Breath Volume		1.992	06:12
Air Blank		0.000	06:12
Air Blank		0.000	06:14
Subject Sample	# 2	0.000	06:15
Breath Volume		3.839	06:15
Air Blank		0.000	06:15
Control Test		0.079	06:16
Air Blank		0.000	06:16
Diagnostics Check		OK	06:16

Data Download Date/Time: 09/19/2016 06:25

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILSON,CHRISTOPHER,R
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: WILSON,CHRISTOPHER
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 08/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A3
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 14:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:41

SUBJECT'S LAST NAME: SCHMITZ

DATE/TIME OF TEST: 09/10/2016 15:09:17

Diagnostics Check		OK	15:06	
Air Blank		0.000	15:06	
Control Test		0.079	15:06	
Air Blank		0.000	15:07	
Subject Sample	# 1		15:07	Subject Test Refused
Breath Volume		0.000	15:07	
Air Blank		0.000	15:08	
Control Test		0.078	15:08	
Air Blank		0.000	15:09	
Diagnostics Check		OK	15:09	

Data Download Date/Time: 09/19/2016 06:25

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 08/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A3
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 18:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:19

SUBJECT'S LAST NAME: CUMMINS

DATE/TIME OF TEST: 09/10/2016 19:45:26

Diagnostics Check		OK	19:41	
Air Blank		0.000	19:42	
Control Test		0.078	19:42	
Air Blank		0.000	19:42	
Subject Sample	# 1		19:44	Subject Test Refused
Breath Volume		0.000	19:44	
Air Blank		0.000	19:44	
Control Test		0.079	19:44	
Air Blank		0.000	19:45	
Diagnostics Check		OK	19:45	

Data Download Date/Time: 09/19/2016 06:25

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 08/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A3
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:49

SUBJECT'S LAST NAME: CURRY

DATE/TIME OF TEST: 09/11/2016 03:18:00

Diagnostics Check		OK	03:10
Air Blank		0.000	03:10
Control Test		0.079	03:11
Air Blank		0.000	03:11
Subject Sample	# 1	0.131	03:13
Breath Volume		2.062	03:13
Air Blank		0.000	03:14
Air Blank		0.000	03:15
Subject Sample	# 2	0.139	03:16
Breath Volume		1.988	03:16
Air Blank		0.000	03:17
Control Test		0.079	03:17
Air Blank		0.000	03:17
Diagnostics Check		OK	03:18

Data Download Date/Time: 09/19/2016 06:26

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FREUDENTHAL,RYAN,T
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: GOSIER,MATT
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 08/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A3
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:58

SUBJECT'S LAST NAME: CARROLL

DATE/TIME OF TEST: 09/12/2016 03:30:12

Diagnostics Check		OK	03:18	
Air Blank		0.000	03:19	
Control Test		0.078	03:19	
Air Blank		0.000	03:19	
Subject Sample	# 1	0.086	03:23	Volume Not Met
Breath Volume		0.695	03:23	
Air Blank		0.000	03:23	
Air Blank		0.000	03:25	
Subject Sample	# 2	0.085	03:28	Volume Not Met
Breath Volume		1.125	03:28	
Air Blank		0.000	03:29	
Control Test		0.076	03:29	
Air Blank		0.000	03:30	
Diagnostics Check		OK	03:30	

Data Download Date/Time: 09/19/2016 06:26

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FREUDENTHAL,RYAN,T
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: GOSIER,MATT
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 08/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A3
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:58

SUBJECT'S LAST NAME: CARROLL

DATE/TIME OF TEST: 09/12/2016 03:41:41

Diagnostics Check		OK	03:34
Air Blank		0.000	03:34
Control Test		0.077	03:35
Air Blank		0.000	03:35
Subject Sample	# 1	0.098	03:37
Breath Volume		1.191	03:37
Air Blank		0.000	03:37
Air Blank		0.000	03:39
Subject Sample	# 2	0.106	03:40
Breath Volume		1.226	03:40
Air Blank		0.000	03:40
Control Test		0.077	03:41
Air Blank		0.000	03:41
Diagnostics Check		OK	03:41

Data Download Date/Time: 09/19/2016 06:26

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUNT,JERRY,D
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: HUNT,JERRY
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 08/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A3
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 13:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:24

SUBJECT'S LAST NAME: PACI

DATE/TIME OF TEST: 09/13/2016 13:53:19

Diagnostics Check		OK	13:44	
Air Blank		0.000	13:45	
Control Test		0.078	13:45	
Air Blank		0.000	13:46	
Subject Sample	# 1	0.000	13:49	Volume Not Met
Breath Volume		0.968	13:49	
Air Blank		0.000	13:49	
Air Blank		0.000	13:51	
Subject Sample	# 2		13:51	Subject Test Refused
Breath Volume		0.000	13:51	
Air Blank		0.000	13:52	
Control Test		0.079	13:52	
Air Blank		0.000	13:53	
Diagnostics Check		OK	13:53	

Data Download Date/Time: 09/19/2016 06:26

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUNT,JERRY,D
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: ASBURY,CHRISTOPHER
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 08/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A3
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 13:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:14

SUBJECT'S LAST NAME: PALMER

DATE/TIME OF TEST: 09/16/2016 14:41:36

Diagnostics Check		OK	14:35
Air Blank		0.000	14:35
Control Test		0.078	14:36
Air Blank		0.000	14:36
Subject Sample	# 1	0.000	14:37
Breath Volume		3.628	14:37
Air Blank		0.000	14:37
Air Blank		0.000	14:39
Subject Sample	# 2	0.000	14:40
Breath Volume		3.542	14:40
Air Blank		0.000	14:40
Control Test		0.079	14:41
Air Blank		0.000	14:41
Diagnostics Check		OK	14:41

Data Download Date/Time: 09/19/2016 06:26

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FREUDENTHAL,RYAN,T
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: GOSIER,MATT
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 08/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A3
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:04

SUBJECT'S LAST NAME: MEEKS

DATE/TIME OF TEST: 09/17/2016 03:37:38

Diagnostics Check		OK	03:28	
Air Blank		0.000	03:28	
Control Test		0.079	03:28	
Air Blank		0.000	03:29	
Subject Sample	# 1		03:32	No Sample Provided
Breath Volume		0.000	03:32	
Air Blank		0.000	03:33	
Air Blank		0.000	03:35	
Subject Sample	# 2		03:36	Subject Test Refused
Breath Volume		0.000	03:36	
Air Blank		0.000	03:36	
Control Test		0.080	03:37	
Air Blank		0.000	03:37	
Diagnostics Check		OK	03:37	

Data Download Date/Time: 09/19/2016 06:26

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ISKRA,KURT,J
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: BOWLING,KEITH
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 08/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A3
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 12:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:29

SUBJECT'S LAST NAME: ADLEMAN

DATE/TIME OF TEST: 09/18/2016 13:04:11

Diagnostics Check		OK	12:50	
Air Blank		0.000	12:51	
Control Test		0.079	12:51	
Air Blank		0.000	12:52	
Subject Sample	# 1	0.144	12:54	
Breath Volume		2.136	12:54	
Air Blank		0.000	12:55	
Air Blank		0.000	12:56	
Subject Sample	# 2	0.131	13:00	Volume Not Met
Breath Volume		0.050	13:00	
Air Blank		0.000	13:00	
Air Blank		0.000	13:02	
Subject Sample	# 3		13:02	Subject Test Refused
Breath Volume		0.000	13:02	
Air Blank		0.000	13:03	
Control Test		0.077	13:03	
Air Blank		0.000	13:04	
Diagnostics Check		OK	13:04	

Data Download Date/Time: 09/19/2016 06:26

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TINDALL,RICHARD,P
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: GOSIER,MATTHEW
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 08/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A3
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:03

SUBJECT'S LAST NAME: MARRA

DATE/TIME OF TEST: 09/19/2016 04:30:35

Diagnostics Check		OK	04:23
Air Blank		0.000	04:24
Control Test		0.079	04:24
Air Blank		0.000	04:25
Subject Sample	# 1	0.148	04:26
Breath Volume		2.015	04:26
Air Blank		0.000	04:26
Air Blank		0.000	04:28
Subject Sample	# 2	0.151	04:29
Breath Volume		1.796	04:29
Air Blank		0.000	04:29
Control Test		0.079	04:29
Air Blank		0.000	04:30
Diagnostics Check		OK	04:30

Data Download Date/Time: 09/19/2016 06:26

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALAMEDA,JORGE,L
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: ALAMEDA,JORGE
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 09/27/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A3
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 15:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:16

SUBJECT'S LAST NAME: CORTEZ

DATE/TIME OF TEST: 09/30/2016 17:57:00

Diagnostics Check		OK	17:50
Air Blank		0.000	17:51
Control Test		0.078	17:51
Air Blank		0.000	17:52
Subject Sample	# 1	0.248	17:52
Breath Volume		2.800	17:52
Air Blank		0.000	17:53
Air Blank		0.000	17:55
Subject Sample	# 2	0.247	17:55
Breath Volume		2.425	17:55
Air Blank		0.000	17:56
Control Test		0.078	17:56
Air Blank		0.000	17:56
Diagnostics Check		OK	17:57

Data Download Date/Time: 10/13/2016 11:44

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FREUDENTHAL,RYAN,T
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: GOSIER,MATT
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 09/27/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A3
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:55

SUBJECT'S LAST NAME: MARINO

DATE/TIME OF TEST: 09/30/2016 21:22:10

Diagnostics Check		OK	21:15
Air Blank		0.000	21:16
Control Test		0.079	21:16
Air Blank		0.000	21:17
Subject Sample	# 1	0.098	21:17
Breath Volume		1.781	21:17
Air Blank		0.000	21:18
Air Blank		0.000	21:20
Subject Sample	# 2	0.101	21:20
Breath Volume		1.550	21:20
Air Blank		0.000	21:21
Control Test		0.078	21:21
Air Blank		0.000	21:22
Diagnostics Check		OK	21:22

Data Download Date/Time: 10/13/2016 11:44

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAWKINS,KIMBERLY,J
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: HOLLINGSWORTH,ERICA
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 08/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 07:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:50

SUBJECT'S LAST NAME: IRVINE

DATE/TIME OF TEST: 09/04/2016 08:37:56

Diagnostics Check		OK	08:35	
Air Blank		0.000	08:35	
Control Test		0.077	08:35	
Air Blank		0.000	08:36	
Subject Sample	# 1		08:36	Subject Test Refused
Breath Volume		0.000	08:36	
Air Blank		0.000	08:36	
Control Test		0.079	08:37	
Air Blank		0.000	08:37	
Diagnostics Check		OK	08:37	

Data Download Date/Time: 09/06/2016 12:03

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRACKLEY,RONALD,L
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: BRACKLEY,RONALD
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 08/05/2016

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 11:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:14

SUBJECT'S LAST NAME: DEL VALLE

DATE/TIME OF TEST: 09/05/2016 11:49:21

Diagnostics Check		OK	11:42
Air Blank		0.000	11:43
Control Test		0.078	11:43
Air Blank		0.000	11:44
Subject Sample	# 1	0.000	11:45
Breath Volume		2.007	11:45
Air Blank		0.000	11:45
Air Blank		0.000	11:47
Subject Sample	# 2	0.000	11:47
Breath Volume		1.683	11:47
Air Blank		0.000	11:48
Control Test		0.078	11:48
Air Blank		0.000	11:49
Diagnostics Check		OK	11:49

Data Download Date/Time: 09/06/2016 12:03

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MELTZER,CHRISTOPHER,J
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: MELTZER,CHRISTOPHER
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 09/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:48

SUBJECT'S LAST NAME: FIGUEROA VARGAS

DATE/TIME OF TEST: 09/16/2016 21:18:22

Diagnostics Check		OK	21:09
Air Blank		0.000	21:10
Control Test		0.078	21:10
Air Blank		0.000	21:11
Subject Sample	# 1	0.123	21:12
Breath Volume		1.460	21:12
Air Blank		0.000	21:13
Air Blank		0.000	21:15
Subject Sample	# 2	0.121	21:16
Breath Volume		1.507	21:16
Air Blank		0.000	21:17
Control Test		0.077	21:17
Air Blank		0.000	21:18
Diagnostics Check		OK	21:18

Data Download Date/Time: 10/06/2016 11:06

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORTIZ,BILLY,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: HAMILTON,KEVIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 08/19/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A1
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 11:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:30

SUBJECT'S LAST NAME: HORNE

DATE/TIME OF TEST: 09/02/2016 12:53:55

Diagnostics Check		OK	12:51	
Air Blank		0.000	12:51	
Control Test		0.077	12:51	
Air Blank		0.000	12:52	
Subject Sample	# 1		12:52	Subject Test Refused
Breath Volume		0.000	12:52	
Air Blank		0.000	12:53	
Control Test		0.077	12:53	
Air Blank		0.000	12:53	
Diagnostics Check		OK	12:53	

Data Download Date/Time: 09/15/2016 10:15

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNORR,PETER,J
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: KNORR,PETER
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 08/19/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A1
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:11

SUBJECT'S LAST NAME: AUSTIN

DATE/TIME OF TEST: 09/03/2016 03:41:21

Diagnostics Check		OK	03:34
Air Blank		0.000	03:34
Control Test		0.079	03:35
Air Blank		0.000	03:35
Subject Sample	# 1	0.027	03:36
Breath Volume		2.554	03:36
Air Blank		0.000	03:37
Air Blank		0.000	03:39
Subject Sample	# 2	0.027	03:39
Breath Volume		3.058	03:39
Air Blank		0.000	03:40
Control Test		0.079	03:40
Air Blank		0.000	03:41
Diagnostics Check		OK	03:41

Data Download Date/Time: 09/15/2016 10:15

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORTIZ,BILLY,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: HAMILTON,KEVIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 08/19/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A1
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 14:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:28

SUBJECT'S LAST NAME: BLANTON

DATE/TIME OF TEST: 09/03/2016 14:51:58

Diagnostics Check		OK	14:49	
Air Blank		0.000	14:49	
Control Test		0.078	14:49	
Air Blank		0.000	14:50	
Subject Sample	# 1		14:50	Subject Test Refused
Breath Volume		0.000	14:50	
Air Blank		0.000	14:51	
Control Test		0.077	14:51	
Air Blank		0.000	14:51	
Diagnostics Check		OK	14:51	

Data Download Date/Time: 09/15/2016 10:15

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NIXON,BARRY,M
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: FITSCHEN,BRYAN
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 08/19/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A1
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 09/04/2016 03:18:41

Diagnostics Check		OK	03:10
Air Blank		0.000	03:11
Control Test		0.079	03:11
Air Blank		0.000	03:12
Subject Sample	# 1	0.141	03:12
Breath Volume		1.367	03:12
Air Blank		0.000	03:13
Air Blank		0.000	03:15
Subject Sample	# 2	0.134	03:17
Breath Volume		1.769	03:17
Air Blank		0.000	03:17
Control Test		0.075	03:18
Air Blank		0.000	03:18
Diagnostics Check		OK	03:18

Data Download Date/Time: 09/15/2016 10:15

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KELLY,WESLEY,R
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: KELLY,WESLEY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 08/19/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A1
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:58

SUBJECT'S LAST NAME: GROSS

DATE/TIME OF TEST: 09/04/2016 04:26:15

Diagnostics Check		OK	04:18
Air Blank		0.000	04:19
Control Test		0.077	04:19
Air Blank		0.000	04:19
Subject Sample	# 1	0.235	04:20
Breath Volume		1.925	04:20
Air Blank		0.000	04:21
Air Blank		0.000	04:23
Subject Sample	# 2	0.252	04:24
Breath Volume		2.511	04:24
Air Blank		0.000	04:25
Control Test		0.079	04:25
Air Blank		0.000	04:26
Diagnostics Check		OK	04:26

Data Download Date/Time: 09/15/2016 10:15

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NOTTINGHAM,CHRISTOPHER,R
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: NOTTINGHAM,CHRISTOPHER
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 08/19/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A1
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:05

SUBJECT'S LAST NAME: ARMSTRONG

DATE/TIME OF TEST: 09/04/2016 05:31:25

Diagnostics Check		OK	05:25
Air Blank		0.000	05:25
Control Test		0.078	05:26
Air Blank		0.000	05:26
Subject Sample	# 1	0.103	05:26
Breath Volume		1.398	05:26
Air Blank		0.000	05:27
Air Blank		0.000	05:29
Subject Sample	# 2	0.103	05:29
Breath Volume		1.777	05:29
Air Blank		0.000	05:30
Control Test		0.080	05:30
Air Blank		0.000	05:31
Diagnostics Check		OK	05:31

Data Download Date/Time: 09/15/2016 10:15

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CINELLI,ADAM,D
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: CINELLI,ADAM
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 08/19/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A1
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:46

SUBJECT'S LAST NAME: STEARNS

DATE/TIME OF TEST: 09/05/2016 02:13:27

Diagnostics Check		OK	02:07
Air Blank		0.000	02:07
Control Test		0.078	02:08
Air Blank		0.000	02:08
Subject Sample	# 1	0.096	02:09
Breath Volume		2.664	02:09
Air Blank		0.000	02:09
Air Blank		0.000	02:11
Subject Sample	# 2	0.095	02:11
Breath Volume		2.746	02:11
Air Blank		0.000	02:12
Control Test		0.078	02:12
Air Blank		0.000	02:13
Diagnostics Check		OK	02:13

Data Download Date/Time: 09/15/2016 10:15

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BECKMAN,KEVIN,E
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)
ARRESTING OFFICER: WILLIAMS,CLINT
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 08/19/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A1
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: BUI
ARREST TIME: 18:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:18

SUBJECT'S LAST NAME: SAWYER

DATE/TIME OF TEST: 09/05/2016 19:42:23

Diagnostics Check		OK	19:38	
Air Blank		0.000	19:39	
Control Test		0.078	19:39	
Air Blank		0.000	19:39	
Subject Sample	# 1		19:41	Subject Test Refused
Breath Volume		0.000	19:41	
Air Blank		0.000	19:41	
Control Test		0.078	19:41	
Air Blank		0.000	19:42	
Diagnostics Check		OK	19:42	

Data Download Date/Time: 09/15/2016 10:15

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BECKMAN,KEVIN,E
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)
ARRESTING OFFICER: BECKMAN,KEVIN
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 08/19/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A1
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:54

SUBJECT'S LAST NAME: BILLINGS

DATE/TIME OF TEST: 09/06/2016 01:21:02

Diagnostics Check		OK	01:14
Air Blank		0.000	01:15
Control Test		0.077	01:15
Air Blank		0.000	01:15
Subject Sample	# 1	0.140	01:16
Breath Volume		1.175	01:16
Air Blank		0.000	01:17
Air Blank		0.000	01:18
Subject Sample	# 2	0.158	01:19
Breath Volume		1.628	01:19
Air Blank		0.000	01:20
Control Test		0.078	01:20
Air Blank		0.000	01:20
Diagnostics Check		OK	01:21

Data Download Date/Time: 09/15/2016 10:15

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AIKEN,DURRIEL,R
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: AIKEN,DURRIEL
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 08/19/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A1
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: MARSH

DATE/TIME OF TEST: 09/06/2016 02:59:15

Diagnostics Check		OK	02:53
Air Blank		0.000	02:53
Control Test		0.078	02:53
Air Blank		0.000	02:54
Subject Sample	# 1	0.000	02:54
Breath Volume		2.574	02:54
Air Blank		0.000	02:55
Air Blank		0.000	02:57
Subject Sample	# 2	0.000	02:57
Breath Volume		2.199	02:57
Air Blank		0.000	02:58
Control Test		0.078	02:58
Air Blank		0.000	02:59
Diagnostics Check		OK	02:59

Data Download Date/Time: 09/15/2016 10:15

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLYTH,JACOB,A
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: BLYTH,JACOB
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 08/19/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A1
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: DOUGHERTY

DATE/TIME OF TEST: 09/08/2016 22:48:07

Diagnostics Check		OK	22:41
Air Blank		0.000	22:41
Control Test		0.079	22:42
Air Blank		0.000	22:42
Subject Sample	# 1	0.110	22:43
Breath Volume		1.445	22:43
Air Blank		0.000	22:44
Air Blank		0.000	22:45
Subject Sample	# 2	0.111	22:46
Breath Volume		1.335	22:46
Air Blank		0.000	22:47
Control Test		0.078	22:47
Air Blank		0.000	22:48
Diagnostics Check		OK	22:48

Data Download Date/Time: 09/15/2016 10:15

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NIXON,BARRY,M
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: NIXON,BARRY
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 08/19/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A1
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: STRIANO

DATE/TIME OF TEST: 09/10/2016 03:38:23

Diagnostics Check		OK	03:30
Air Blank		0.000	03:31
Control Test		0.078	03:31
Air Blank		0.000	03:32
Subject Sample	# 1	0.222	03:32
Breath Volume		1.878	03:32
Air Blank		0.000	03:33
Air Blank		0.000	03:35
Subject Sample	# 2	0.223	03:36
Breath Volume		2.199	03:36
Air Blank		0.000	03:37
Control Test		0.078	03:37
Air Blank		0.000	03:38
Diagnostics Check		OK	03:38

Data Download Date/Time: 09/15/2016 10:15

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASTILLO,JOSUET,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: CASTILLO,JOSUET
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 08/19/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A1
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: GOMEZ

DATE/TIME OF TEST: 09/10/2016 23:41:01

Diagnostics Check		OK	23:37	
Air Blank		0.000	23:37	
Control Test		0.078	23:37	
Air Blank		0.000	23:38	
Subject Sample	# 1		23:39	Subject Test Refused
Breath Volume		0.000	23:39	
Air Blank		0.000	23:40	
Control Test		0.078	23:40	
Air Blank		0.000	23:40	
Diagnostics Check		OK	23:41	

Data Download Date/Time: 09/15/2016 10:15

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,WILLIAM,T
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SMITH,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 08/19/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A1
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:16

SUBJECT'S LAST NAME: BURROWS

DATE/TIME OF TEST: 09/11/2016 00:43:04

Diagnostics Check		OK	00:36
Air Blank		0.000	00:37
Control Test		0.080	00:37
Air Blank		0.000	00:38
Subject Sample	# 1	0.000	00:38
Breath Volume		1.804	00:38
Air Blank		0.000	00:39
Air Blank		0.000	00:41
Subject Sample	# 2	0.000	00:41
Breath Volume		1.753	00:41
Air Blank		0.000	00:42
Control Test		0.078	00:42
Air Blank		0.000	00:42
Diagnostics Check		OK	00:43

Data Download Date/Time: 09/15/2016 10:15

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NIXON,BARRY,M
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: NIXON,BARRY
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 08/19/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A1
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:17

SUBJECT'S LAST NAME: LUGO GONZALEZ

DATE/TIME OF TEST: 09/11/2016 03:45:26

Diagnostics Check		OK	03:38
Air Blank		0.000	03:38
Control Test		0.078	03:38
Air Blank		0.000	03:39
Subject Sample	# 1	0.134	03:40
Breath Volume		1.410	03:40
Air Blank		0.000	03:40
Air Blank		0.000	03:42
Subject Sample	# 2	0.129	03:43
Breath Volume		1.542	03:43
Air Blank		0.000	03:44
Control Test		0.078	03:44
Air Blank		0.000	03:45
Diagnostics Check		OK	03:45

Data Download Date/Time: 09/15/2016 10:15

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASTILLO,JOSUET,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: CASTILLO,JOSUET
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 08/19/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A1
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 19:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:45

SUBJECT'S LAST NAME: DUNCAN

DATE/TIME OF TEST: 09/14/2016 21:12:08

Diagnostics Check		OK	21:05
Air Blank		0.000	21:06
Control Test		0.078	21:06
Air Blank		0.000	21:06
Subject Sample	# 1	0.112	21:07
Breath Volume		1.351	21:07
Air Blank		0.000	21:07
Air Blank		0.000	21:09
Subject Sample	# 2	0.111	21:10
Breath Volume		1.214	21:10
Air Blank		0.000	21:11
Control Test		0.079	21:11
Air Blank		0.000	21:12
Diagnostics Check		OK	21:12

Data Download Date/Time: 09/15/2016 10:15

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AIKEN,DURRIEL,R
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: AIKEN,DURRIEL
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 08/19/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A1
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: PALMER

DATE/TIME OF TEST: 09/15/2016 00:09:42

Diagnostics Check		OK	23:56	
Air Blank		0.000	23:57	
Control Test		0.077	23:57	
Air Blank		0.000	23:57	
Subject Sample	# 1		00:01	No Sample Provided
Breath Volume		0.000	00:01	
Air Blank		0.000	00:01	
Air Blank		0.000	00:03	
Subject Sample	# 2	0.199	00:03	
Breath Volume		1.234	00:03	
Air Blank		0.000	00:04	
Air Blank		0.000	00:06	
Subject Sample	# 3	0.188	00:08	
Breath Volume		1.640	00:08	
Air Blank		0.000	00:08	
Control Test		0.077	00:09	
Air Blank		0.000	00:09	
Diagnostics Check		OK	00:09	

Data Download Date/Time: 09/15/2016 10:15

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,WILLIAM,T
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SMITH,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 09/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A1
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: KEARNS

DATE/TIME OF TEST: 09/16/2016 02:14:16

Diagnostics Check		OK	02:07
Air Blank		0.000	02:07
Control Test		0.079	02:07
Air Blank		0.000	02:08
Subject Sample	# 1	0.000	02:09
Breath Volume		1.261	02:09
Air Blank		0.000	02:09
Air Blank		0.000	02:11
Subject Sample	# 2	0.000	02:12
Breath Volume		1.257	02:12
Air Blank		0.000	02:13
Control Test		0.079	02:13
Air Blank		0.000	02:14
Diagnostics Check		OK	02:14

Data Download Date/Time: 10/14/2016 08:29

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALVARADO,ANGEL,L
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: ALVARADO,ANGEL
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 09/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A1
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: JOHNES

DATE/TIME OF TEST: 09/16/2016 04:18:01

Diagnostics Check		OK	04:11
Air Blank		0.000	04:12
Control Test		0.078	04:12
Air Blank		0.000	04:12
Subject Sample	# 1	0.119	04:13
Breath Volume		1.789	04:13
Air Blank		0.000	04:14
Air Blank		0.000	04:16
Subject Sample	# 2	0.121	04:16
Breath Volume		1.414	04:16
Air Blank		0.000	04:17
Control Test		0.078	04:17
Air Blank		0.000	04:17
Diagnostics Check		OK	04:18

Data Download Date/Time: 10/14/2016 08:29

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNORR,PETER,J
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: KNORR,PETER
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 09/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A1
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: KIRKLAND

DATE/TIME OF TEST: 09/17/2016 00:28:42

Diagnostics Check		OK	00:26	
Air Blank		0.000	00:27	
Control Test		0.078	00:27	
Air Blank		0.000	00:27	
Subject Sample	# 1		00:28	RFI Detect
Breath Volume		0.000	00:28	
Air Blank			00:28	RFI Detect
Air Blank		0.000	00:28	

Data Download Date/Time: 10/14/2016 08:29

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNORR,PETER,J
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: KNORR,PETER
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 09/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A1
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: KIRKLAND

DATE/TIME OF TEST: 09/17/2016 00:37:00

Diagnostics Check		OK	00:30
Air Blank		0.000	00:31
Control Test		0.078	00:31
Air Blank		0.000	00:31
Subject Sample	# 1	0.175	00:32
Breath Volume		2.929	00:32
Air Blank		0.000	00:33
Air Blank		0.000	00:34
Subject Sample	# 2	0.171	00:35
Breath Volume		2.890	00:35
Air Blank		0.000	00:36
Control Test		0.078	00:36
Air Blank		0.000	00:36
Diagnostics Check		OK	00:37

Data Download Date/Time: 10/14/2016 08:29

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNORR,PETER,J
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: BLICKER,E
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 09/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A1
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:03

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 09/17/2016 03:28:29

Diagnostics Check		OK	03:24	
Air Blank		0.000	03:25	
Control Test		0.077	03:25	
Air Blank		0.000	03:26	
Subject Sample	# 1		03:27	Subject Test Refused
Breath Volume		0.000	03:27	
Air Blank		0.000	03:27	
Control Test		0.078	03:27	
Air Blank		0.000	03:28	
Diagnostics Check		OK	03:28	

Data Download Date/Time: 10/14/2016 08:29

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,NELSON,X
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: RODRIGUEZ,NELSON
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 09/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A1
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: THIBODEAU

DATE/TIME OF TEST: 09/17/2016 04:05:25

Diagnostics Check		OK	04:02	
Air Blank		0.000	04:03	
Control Test		0.079	04:03	
Air Blank		0.000	04:03	
Subject Sample	# 1		04:04	Subject Test Refused
Breath Volume		0.000	04:04	
Air Blank		0.000	04:04	
Control Test		0.078	04:04	
Air Blank		0.000	04:05	
Diagnostics Check		OK	04:05	

Data Download Date/Time: 10/14/2016 08:29

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONSEGUR,ROBERT,
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: FITSCHEN,BRYAN
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 09/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A1
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:58

SUBJECT'S LAST NAME: COGGEN

DATE/TIME OF TEST: 09/18/2016 03:25:04

Diagnostics Check		OK	03:18
Air Blank		0.000	03:18
Control Test		0.078	03:18
Air Blank		0.000	03:19
Subject Sample	# 1	0.158	03:20
Breath Volume		1.437	03:20
Air Blank		0.000	03:20
Air Blank		0.000	03:22
Subject Sample	# 2	0.167	03:23
Breath Volume		2.796	03:23
Air Blank		0.000	03:24
Control Test		0.078	03:24
Air Blank		0.000	03:24
Diagnostics Check		OK	03:25

Data Download Date/Time: 10/14/2016 08:29

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,WILLIAM,T
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SMITH,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 09/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A1
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:22

SUBJECT'S LAST NAME: KING

DATE/TIME OF TEST: 09/20/2016 23:49:29

Diagnostics Check		OK	23:42
Air Blank		0.000	23:42
Control Test		0.079	23:43
Air Blank		0.000	23:43
Subject Sample	# 1	0.075	23:44
Breath Volume		1.218	23:44
Air Blank		0.000	23:45
Air Blank		0.000	23:47
Subject Sample	# 2	0.075	23:48
Breath Volume		1.605	23:48
Air Blank		0.000	23:48
Control Test		0.078	23:48
Air Blank		0.000	23:49
Diagnostics Check		OK	23:49

Data Download Date/Time: 10/14/2016 08:29

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KELLY,WESLEY,R
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: KELLY,WESLEY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 09/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A1
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:37

SUBJECT'S LAST NAME: SANCHEZ

DATE/TIME OF TEST: 09/22/2016 04:03:32

Diagnostics Check		OK	03:57
Air Blank		0.000	03:57
Control Test		0.079	03:58
Air Blank		0.000	03:58
Subject Sample	# 1	0.132	03:59
Breath Volume		2.593	03:59
Air Blank		0.000	03:59
Air Blank		0.000	04:01
Subject Sample	# 2	0.135	04:02
Breath Volume		3.511	04:02
Air Blank		0.000	04:02
Control Test		0.079	04:02
Air Blank		0.000	04:03
Diagnostics Check		OK	04:03

Data Download Date/Time: 10/14/2016 08:29

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NOTTINGHAM,CHRISTOPHER,R
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: NOTTINGHAM,CHRISTOPHER
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 09/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A1
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:12

SUBJECT'S LAST NAME: TANSY

DATE/TIME OF TEST: 09/23/2016 03:45:53

Diagnostics Check		OK	03:32	
Air Blank		0.000	03:32	
Control Test		0.080	03:33	
Air Blank		0.000	03:33	
Subject Sample	# 1	0.193	03:36	Slope Not Level
Breath Volume		1.730	03:36	
Air Blank		0.000	03:37	
Air Blank		0.000	03:39	
Subject Sample	# 2	0.198	03:41	
Breath Volume		1.945	03:41	
Air Blank		0.000	03:42	
Air Blank		0.000	03:43	
Subject Sample	# 3	0.208	03:44	
Breath Volume		1.621	03:44	
Air Blank		0.000	03:44	
Control Test		0.077	03:45	
Air Blank		0.000	03:45	
Diagnostics Check		OK	03:45	

Data Download Date/Time: 10/14/2016 08:29

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NIXON,BARRY,M
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: NIXON,BARRY
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 09/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A1
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:48

SUBJECT'S LAST NAME: FRENCH

DATE/TIME OF TEST: 09/23/2016 21:15:00

Diagnostics Check		OK	21:08
Air Blank		0.000	21:09
Control Test		0.078	21:09
Air Blank		0.000	21:10
Subject Sample	# 1	0.175	21:10
Breath Volume		2.605	21:10
Air Blank		0.000	21:11
Air Blank		0.000	21:13
Subject Sample	# 2	0.187	21:13
Breath Volume		2.261	21:13
Air Blank		0.000	21:14
Control Test		0.078	21:14
Air Blank		0.000	21:14
Diagnostics Check		OK	21:15

Data Download Date/Time: 10/14/2016 08:29

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NIXON,BARRY,
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: NIXON,BARRY
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 09/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A1
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:52

SUBJECT'S LAST NAME: OVERSTREET

DATE/TIME OF TEST: 09/24/2016 03:15:26

Diagnostics Check		OK	03:12	
Air Blank		0.000	03:12	
Control Test		0.078	03:13	
Air Blank		0.000	03:13	
Subject Sample	# 1		03:14	Subject Test Refused
Breath Volume		0.000	03:14	
Air Blank		0.000	03:14	
Control Test		0.078	03:14	
Air Blank		0.000	03:15	
Diagnostics Check		OK	03:15	

Data Download Date/Time: 10/14/2016 08:29

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRIOLAT,CASEY,V
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: BRIOLAT,CASEY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 09/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A1
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: BAGLANZIS

DATE/TIME OF TEST: 09/25/2016 00:28:40

Diagnostics Check		OK	00:25	
Air Blank		0.000	00:25	
Control Test		0.078	00:26	
Air Blank		0.000	00:26	
Subject Sample	# 1		00:27	Subject Test Refused
Breath Volume		0.000	00:27	
Air Blank		0.000	00:27	
Control Test		0.078	00:28	
Air Blank		0.000	00:28	
Diagnostics Check		OK	00:28	

Data Download Date/Time: 10/14/2016 08:29

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROWAND,ART,
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: PENNELL,RUSSELL
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 09/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A1
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: BOOCHER

DATE/TIME OF TEST: 09/25/2016 02:21:23

Diagnostics Check		OK	02:15
Air Blank		0.000	02:15
Control Test		0.078	02:16
Air Blank		0.000	02:16
Subject Sample	# 1	0.168	02:16
Breath Volume		1.843	02:16
Air Blank		0.000	02:17
Air Blank		0.000	02:19
Subject Sample	# 2	0.170	02:19
Breath Volume		1.121	02:19
Air Blank		0.000	02:20
Control Test		0.078	02:20
Air Blank		0.000	02:21
Diagnostics Check		OK	02:21

Data Download Date/Time: 10/14/2016 08:29

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KELLY,WESLEY,R
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: KELLY,WESLEY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 09/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A1
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:56

SUBJECT'S LAST NAME: TAPIA

DATE/TIME OF TEST: 09/27/2016 22:23:32

Diagnostics Check		OK	22:16
Air Blank		0.000	22:16
Control Test		0.079	22:17
Air Blank		0.000	22:17
Subject Sample	# 1	0.181	22:18
Breath Volume		1.429	22:18
Air Blank		0.000	22:18
Air Blank		0.000	22:20
Subject Sample	# 2	0.181	22:22
Breath Volume		1.796	22:22
Air Blank		0.000	22:22
Control Test		0.081	22:22
Air Blank		0.000	22:23
Diagnostics Check		OK	22:23

Data Download Date/Time: 10/14/2016 08:29

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CINELLI,ADAM,D
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: CINELLI,ADAM
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 09/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A1
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 16:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:34

SUBJECT'S LAST NAME: DACCHILLE

DATE/TIME OF TEST: 09/28/2016 18:16:42

Diagnostics Check		OK	18:10
Air Blank		0.000	18:11
Control Test		0.079	18:11
Air Blank		0.000	18:12
Subject Sample	# 1	0.000	18:12
Breath Volume		2.828	18:12
Air Blank		0.000	18:13
Air Blank		0.000	18:15
Subject Sample	# 2	0.000	18:15
Breath Volume		1.988	18:15
Air Blank		0.000	18:15
Control Test		0.078	18:16
Air Blank		0.000	18:16
Diagnostics Check		OK	18:16

Data Download Date/Time: 10/14/2016 08:29

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNORR,PETER,J
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: KNORR,PETER
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 09/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A1
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: FITZGERALD

DATE/TIME OF TEST: 09/28/2016 23:58:16

Diagnostics Check		OK	23:55	
Air Blank		0.000	23:55	
Control Test		0.080	23:56	
Air Blank		0.000	23:56	
Subject Sample	# 1		23:56	Subject Test Refused
Breath Volume		0.000	23:56	
Air Blank		0.000	23:57	
Control Test		0.078	23:57	
Air Blank		0.000	23:58	
Diagnostics Check		OK	23:58	

Data Download Date/Time: 10/14/2016 08:29

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NIXON,BARRY,
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: NIXON,BARRY
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 09/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A1
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:32

SUBJECT'S LAST NAME: LAW

DATE/TIME OF TEST: 09/29/2016 03:59:27

Diagnostics Check		OK	03:53
Air Blank		0.000	03:54
Control Test		0.079	03:54
Air Blank		0.000	03:54
Subject Sample	# 1	0.000	03:55
Breath Volume		1.328	03:55
Air Blank		0.000	03:55
Air Blank		0.000	03:57
Subject Sample	# 2	0.000	03:58
Breath Volume		1.746	03:58
Air Blank		0.000	03:58
Control Test		0.078	03:58
Air Blank		0.000	03:59
Diagnostics Check		OK	03:59

Data Download Date/Time: 10/14/2016 08:29

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,WILLIAM,T
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SMITH,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 09/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A1
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:19

SUBJECT'S LAST NAME: PALAZZOLO

DATE/TIME OF TEST: 09/29/2016 22:46:25

Diagnostics Check		OK	22:40
Air Blank		0.000	22:40
Control Test		0.079	22:40
Air Blank		0.000	22:41
Subject Sample	# 1	0.067	22:42
Breath Volume		1.832	22:42
Air Blank		0.000	22:42
Air Blank		0.000	22:44
Subject Sample	# 2	0.072	22:44
Breath Volume		1.832	22:44
Air Blank		0.000	22:45
Control Test		0.079	22:45
Air Blank		0.000	22:46
Diagnostics Check		OK	22:46

Data Download Date/Time: 10/14/2016 08:29

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,WILLIM,T
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SMITH,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 09/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A1
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:58

SUBJECT'S LAST NAME: KLIDONAS

DATE/TIME OF TEST: 09/30/2016 05:30:35

Diagnostics Check		OK	05:22
Air Blank		0.000	05:23
Control Test		0.079	05:23
Air Blank		0.000	05:24
Subject Sample	# 1	0.035	05:25
Breath Volume		1.203	05:25
Air Blank		0.000	05:26
Air Blank		0.000	05:28
Subject Sample	# 2	0.033	05:29
Breath Volume		1.316	05:29
Air Blank		0.000	05:29
Control Test		0.078	05:30
Air Blank		0.000	05:30
Diagnostics Check		OK	05:30

Data Download Date/Time: 10/14/2016 08:29

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KRAUS,CHRISTOPHER,J
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: KRAUS,CHRISTOPHER
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 09/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A1
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 12:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 12:20

SUBJECT'S LAST NAME: CROFT

DATE/TIME OF TEST: 09/30/2016 13:02:08

Diagnostics Check		OK	12:56
Air Blank		0.000	12:56
Control Test		0.078	12:56
Air Blank		0.000	12:57
Subject Sample	# 1	0.294	12:57
Breath Volume		1.597	12:57
Air Blank		0.000	12:58
Air Blank		0.000	13:00
Subject Sample	# 2	0.300	13:00
Breath Volume		1.546	13:00
Air Blank		0.000	13:01
Control Test		0.079	13:01
Air Blank		0.000	13:02
Diagnostics Check		OK	13:02

Data Download Date/Time: 10/14/2016 08:29

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: KEEN,C
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 08/31/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:08

SUBJECT'S LAST NAME: CANOVA

DATE/TIME OF TEST: 09/02/2016 04:31:56

Diagnostics Check		OK	04:29	
Air Blank		0.000	04:29	
Control Test		0.078	04:29	
Air Blank		0.000	04:30	
Subject Sample	# 1		04:30	Subject Test Refused
Breath Volume		0.000	04:30	
Air Blank		0.000	04:31	
Control Test		0.076	04:31	
Air Blank		0.000	04:31	
Diagnostics Check		OK	04:31	

Data Download Date/Time: 09/14/2016 13:03

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BARTILUCCI,D
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/31/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:21

SUBJECT'S LAST NAME: AYALA

DATE/TIME OF TEST: 09/03/2016 02:47:45

Diagnostics Check		OK	02:41
Air Blank		0.000	02:41
Control Test		0.077	02:42
Air Blank		0.000	02:42
Subject Sample	# 1	0.177	02:43
Breath Volume		2.410	02:43
Air Blank		0.000	02:43
Air Blank		0.000	02:45
Subject Sample	# 2	0.179	02:46
Breath Volume		1.937	02:46
Air Blank		0.000	02:46
Control Test		0.077	02:47
Air Blank		0.000	02:47
Diagnostics Check		OK	02:47

Data Download Date/Time: 09/14/2016 13:03

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HILLIER,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/31/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:53

SUBJECT'S LAST NAME: SAMUELS

DATE/TIME OF TEST: 09/03/2016 04:20:39

Diagnostics Check		OK	04:14
Air Blank		0.000	04:14
Control Test		0.077	04:15
Air Blank		0.000	04:15
Subject Sample	# 1	0.110	04:16
Breath Volume		2.066	04:16
Air Blank		0.000	04:16
Air Blank		0.000	04:18
Subject Sample	# 2	0.112	04:19
Breath Volume		3.175	04:19
Air Blank		0.000	04:19
Control Test		0.077	04:20
Air Blank		0.000	04:20
Diagnostics Check		OK	04:20

Data Download Date/Time: 09/14/2016 13:03

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: EANES,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/31/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: TORRES BYRNE

DATE/TIME OF TEST: 09/04/2016 03:01:42

Diagnostics Check		OK	02:58	
Air Blank		0.000	02:59	
Control Test		0.077	02:59	
Air Blank		0.000	03:00	
Subject Sample	# 1		03:00	Subject Test Refused
Breath Volume		0.000	03:00	
Air Blank		0.000	03:00	
Control Test		0.077	03:01	
Air Blank		0.000	03:01	
Diagnostics Check		OK	03:01	

Data Download Date/Time: 09/14/2016 13:03

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MARRERO,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/31/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:54

SUBJECT'S LAST NAME: TOTON

DATE/TIME OF TEST: 09/05/2016 00:20:23

Diagnostics Check		OK	00:14
Air Blank		0.000	00:14
Control Test		0.078	00:15
Air Blank		0.000	00:15
Subject Sample	# 1	0.271	00:16
Breath Volume		3.445	00:16
Air Blank		0.000	00:16
Air Blank		0.000	00:18
Subject Sample	# 2	0.276	00:18
Breath Volume		2.871	00:18
Air Blank		0.000	00:19
Control Test		0.077	00:19
Air Blank		0.000	00:20
Diagnostics Check		OK	00:20

Data Download Date/Time: 09/14/2016 13:03

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: VANHOF,J
ARRESTING OFFICER AGENCY: Atlantic Beach PD
LAST AGENCY INSPECTION: 08/31/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:36

SUBJECT'S LAST NAME: LOMBARDI

DATE/TIME OF TEST: 09/05/2016 04:03:08

Diagnostics Check		OK	03:56
Air Blank		0.000	03:57
Control Test		0.077	03:57
Air Blank		0.000	03:57
Subject Sample	# 1	0.127	03:58
Breath Volume		1.785	03:58
Air Blank		0.000	03:59
Air Blank		0.000	04:00
Subject Sample	# 2	0.129	04:01
Breath Volume		1.855	04:01
Air Blank		0.000	04:02
Control Test		0.078	04:02
Air Blank		0.000	04:03
Diagnostics Check		OK	04:03

Data Download Date/Time: 09/14/2016 13:03

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MOORE,A
ARRESTING OFFICER AGENCY: Atlantic Beach PD
LAST AGENCY INSPECTION: 08/31/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:43

SUBJECT'S LAST NAME: FAIRMAN

DATE/TIME OF TEST: 09/07/2016 02:06:40

Diagnostics Check		OK	02:03	
Air Blank		0.000	02:04	
Control Test		0.077	02:04	
Air Blank		0.000	02:05	
Subject Sample	# 1		02:05	Subject Test Refused
Breath Volume		0.000	02:05	
Air Blank		0.000	02:05	
Control Test		0.077	02:06	
Air Blank		0.000	02:06	
Diagnostics Check		OK	02:06	

Data Download Date/Time: 09/14/2016 13:03

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PADGETT,C,N
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHITE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/31/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:58

SUBJECT'S LAST NAME: BIRINCSIK

DATE/TIME OF TEST: 09/08/2016 22:25:40

Diagnostics Check		OK	22:19
Air Blank		0.000	22:19
Control Test		0.077	22:19
Air Blank		0.000	22:20
Subject Sample	# 1	0.273	22:21
Breath Volume		2.996	22:21
Air Blank		0.000	22:21
Air Blank		0.000	22:23
Subject Sample	# 2	0.284	22:24
Breath Volume		2.472	22:24
Air Blank		0.000	22:24
Control Test		0.077	22:25
Air Blank		0.000	22:25
Diagnostics Check		OK	22:25

Data Download Date/Time: 09/14/2016 13:03

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PENNINGTON,EMILY,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: STANLEY,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/31/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:21

SUBJECT'S LAST NAME: EDWARDS

DATE/TIME OF TEST: 09/08/2016 23:52:50

Diagnostics Check		OK	23:49	
Air Blank		0.000	23:50	
Control Test		0.078	23:50	
Air Blank		0.000	23:51	
Subject Sample	# 1		23:51	Subject Test Refused
Breath Volume		0.000	23:51	
Air Blank		0.000	23:51	
Control Test		0.077	23:52	
Air Blank		0.000	23:52	
Diagnostics Check		OK	23:52	

Data Download Date/Time: 09/14/2016 13:03

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PENNINGTON,EMILY,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: GAZIC,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/31/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:46

SUBJECT'S LAST NAME: YATES

DATE/TIME OF TEST: 09/09/2016 04:13:16

Diagnostics Check		OK	04:10	
Air Blank		0.000	04:10	
Control Test		0.077	04:11	
Air Blank		0.000	04:11	
Subject Sample	# 1		04:11	Subject Test Refused
Breath Volume		0.000	04:11	
Air Blank		0.000	04:12	
Control Test		0.077	04:12	
Air Blank		0.000	04:13	
Diagnostics Check		OK	04:13	

Data Download Date/Time: 09/14/2016 13:03

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THIGPEN,J,T
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FISETTE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/31/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:18

SUBJECT'S LAST NAME: TAMPA

DATE/TIME OF TEST: 09/09/2016 22:49:29

Diagnostics Check		OK	22:39
Air Blank		0.000	22:39
Control Test		0.077	22:40
Air Blank		0.000	22:40
Subject Sample	# 1	0.111	22:41
Breath Volume		1.191	22:41
Air Blank		0.000	22:41
Air Blank		0.000	22:43
Subject Sample	# 2	0.132	22:44
Breath Volume		3.738	22:44
Air Blank		0.000	22:45
Air Blank		0.000	22:47
Subject Sample	# 3	0.105	22:47
Breath Volume		1.296	22:47
Air Blank		0.000	22:48
Control Test		0.078	22:48
Air Blank		0.000	22:49
Diagnostics Check		OK	22:49

Data Download Date/Time: 09/14/2016 13:03

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HOOVER,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/31/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:46

SUBJECT'S LAST NAME: MALONE

DATE/TIME OF TEST: 09/10/2016 03:13:49

Diagnostics Check		OK	03:07
Air Blank		0.000	03:07
Control Test		0.078	03:07
Air Blank		0.000	03:08
Subject Sample	# 1	0.122	03:09
Breath Volume		2.839	03:09
Air Blank		0.000	03:09
Air Blank		0.000	03:11
Subject Sample	# 2	0.115	03:12
Breath Volume		1.917	03:12
Air Blank		0.000	03:12
Control Test		0.077	03:13
Air Blank		0.000	03:13
Diagnostics Check		OK	03:13

Data Download Date/Time: 09/14/2016 13:03

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FISETTE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/31/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:07

SUBJECT'S LAST NAME: GEBHART

DATE/TIME OF TEST: 09/10/2016 23:33:45

Diagnostics Check		OK	23:27
Air Blank		0.000	23:27
Control Test		0.076	23:28
Air Blank		0.000	23:28
Subject Sample	# 1	0.270	23:29
Breath Volume		1.582	23:29
Air Blank		0.000	23:30
Air Blank		0.000	23:31
Subject Sample	# 2	0.280	23:32
Breath Volume		1.437	23:32
Air Blank		0.000	23:32
Control Test		0.076	23:33
Air Blank		0.000	23:33
Diagnostics Check		OK	23:33

Data Download Date/Time: 09/14/2016 13:03

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: TAVARES,S
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 08/31/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:31

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 09/11/2016 05:54:05

Diagnostics Check		OK	05:51	
Air Blank		0.000	05:51	
Control Test		0.078	05:52	
Air Blank		0.000	05:52	
Subject Sample	# 1		05:52	Subject Test Refused
Breath Volume		0.000	05:52	
Air Blank		0.000	05:53	
Control Test		0.077	05:53	
Air Blank		0.000	05:54	
Diagnostics Check		OK	05:54	

Data Download Date/Time: 09/14/2016 13:04

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PADGETT,C,N
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: STANLEY,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/31/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:17

SUBJECT'S LAST NAME: COTTON

DATE/TIME OF TEST: 09/11/2016 21:44:13

Diagnostics Check		OK	21:37
Air Blank		0.000	21:38
Control Test		0.078	21:38
Air Blank		0.000	21:39
Subject Sample	# 1	0.000	21:39
Breath Volume		2.636	21:39
Air Blank		0.000	21:40
Air Blank		0.000	21:42
Subject Sample	# 2	0.000	21:42
Breath Volume		2.230	21:42
Air Blank		0.000	21:43
Control Test		0.077	21:43
Air Blank		0.000	21:44
Diagnostics Check		OK	21:44

Data Download Date/Time: 09/14/2016 13:04

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CAMACHO,K
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 08/31/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:26

SUBJECT'S LAST NAME: MCCOY

DATE/TIME OF TEST: 09/11/2016 23:49:06

Diagnostics Check		OK	23:46	
Air Blank		0.000	23:46	
Control Test		0.077	23:47	
Air Blank		0.000	23:47	
Subject Sample	# 1		23:47	Subject Test Refused
Breath Volume		0.000	23:47	
Air Blank		0.000	23:48	
Control Test		0.077	23:48	
Air Blank		0.000	23:49	
Diagnostics Check		OK	23:49	

Data Download Date/Time: 09/14/2016 13:04

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: STANLEY,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/31/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:03

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 09/12/2016 03:30:01

Diagnostics Check		OK	03:23
Air Blank		0.000	03:24
Control Test		0.078	03:24
Air Blank		0.000	03:25
Subject Sample	# 1	0.000	03:25
Breath Volume		1.273	03:25
Air Blank		0.000	03:26
Air Blank		0.000	03:28
Subject Sample	# 2	0.000	03:28
Breath Volume		2.527	03:28
Air Blank		0.000	03:29
Control Test		0.077	03:29
Air Blank		0.000	03:29
Diagnostics Check		OK	03:30

Data Download Date/Time: 09/14/2016 13:04

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LADUE,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/31/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: DIRAR

DATE/TIME OF TEST: 09/12/2016 04:23:00

Diagnostics Check		OK	04:13
Air Blank		0.000	04:14
Control Test		0.076	04:14
Air Blank		0.000	04:14
Subject Sample	# 1	0.177	04:16
Breath Volume		1.378	04:16
Air Blank		0.000	04:17
Air Blank		0.000	04:18
Subject Sample	# 2	0.177	04:21
Breath Volume		1.394	04:21
Air Blank		0.000	04:22
Control Test		0.077	04:22
Air Blank		0.000	04:22
Diagnostics Check		OK	04:23

Data Download Date/Time: 09/14/2016 13:04

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FISSEL,J
ARRESTING OFFICER AGENCY: Atlantic Beach PD
LAST AGENCY INSPECTION: 08/31/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:11

SUBJECT'S LAST NAME: WARDREP

DATE/TIME OF TEST: 09/14/2016 02:37:49

Diagnostics Check		OK	02:31
Air Blank		0.000	02:31
Control Test		0.077	02:32
Air Blank		0.000	02:32
Subject Sample	# 1	0.257	02:33
Breath Volume		3.367	02:33
Air Blank		0.000	02:33
Air Blank		0.000	02:35
Subject Sample	# 2	0.255	02:36
Breath Volume		2.785	02:36
Air Blank		0.000	02:36
Control Test		0.076	02:37
Air Blank		0.000	02:37
Diagnostics Check		OK	02:37

Data Download Date/Time: 09/14/2016 13:04

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FILBERT,E
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 09/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:01

SUBJECT'S LAST NAME: HOFFMAN

DATE/TIME OF TEST: 09/14/2016 23:30:51

Diagnostics Check		OK	23:24
Air Blank		0.000	23:24
Control Test		0.077	23:25
Air Blank		0.000	23:25
Subject Sample	# 1	0.166	23:26
Breath Volume		2.351	23:26
Air Blank		0.000	23:26
Air Blank		0.000	23:28
Subject Sample	# 2	0.185	23:29
Breath Volume		2.238	23:29
Air Blank		0.000	23:29
Control Test		0.076	23:30
Air Blank		0.000	23:30
Diagnostics Check		OK	23:30

Data Download Date/Time: 09/29/2016 12:28

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MOSLEY,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 09/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:49

SUBJECT'S LAST NAME: HANOVER

DATE/TIME OF TEST: 09/15/2016 00:12:51

Diagnostics Check		OK	00:09	
Air Blank		0.000	00:10	
Control Test		0.077	00:10	
Air Blank		0.000	00:11	
Subject Sample	# 1		00:11	Subject Test Refused
Breath Volume		0.000	00:11	
Air Blank		0.000	00:11	
Control Test		0.077	00:12	
Air Blank		0.000	00:12	
Diagnostics Check		OK	00:12	

Data Download Date/Time: 09/29/2016 12:28

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BYRNE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 09/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: NORMAN

DATE/TIME OF TEST: 09/15/2016 03:22:22

Diagnostics Check		OK	03:15
Air Blank		0.000	03:16
Control Test		0.077	03:16
Air Blank		0.000	03:16
Subject Sample	# 1	0.160	03:17
Breath Volume		1.164	03:17
Air Blank		0.000	03:18
Air Blank		0.000	03:20
Subject Sample	# 2	0.158	03:20
Breath Volume		1.304	03:20
Air Blank		0.000	03:21
Control Test		0.077	03:21
Air Blank		0.000	03:22
Diagnostics Check		OK	03:22

Data Download Date/Time: 09/29/2016 12:28

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: KERSHAW,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 09/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:48

SUBJECT'S LAST NAME: MEDINA RODRIGUEZ

DATE/TIME OF TEST: 09/15/2016 04:15:46

Diagnostics Check		OK	04:09
Air Blank		0.000	04:09
Control Test		0.077	04:09
Air Blank		0.000	04:10
Subject Sample	# 1	0.210	04:11
Breath Volume		2.058	04:11
Air Blank		0.000	04:11
Air Blank		0.000	04:13
Subject Sample	# 2	0.204	04:14
Breath Volume		1.503	04:14
Air Blank		0.000	04:14
Control Test		0.076	04:15
Air Blank		0.000	04:15
Diagnostics Check		OK	04:15

Data Download Date/Time: 09/29/2016 12:28

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BROWN,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 09/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:11

SUBJECT'S LAST NAME: POTTHAST

DATE/TIME OF TEST: 09/16/2016 03:36:19

Diagnostics Check		OK	03:32	
Air Blank		0.000	03:33	
Control Test		0.078	03:33	
Air Blank		0.000	03:34	
Subject Sample	# 1		03:34	Subject Test Refused
Breath Volume		0.000	03:34	
Air Blank		0.000	03:35	
Control Test		0.077	03:35	
Air Blank		0.000	03:36	
Diagnostics Check		OK	03:36	

Data Download Date/Time: 09/29/2016 12:28

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BAI,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 09/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:01

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 09/17/2016 00:28:19

Diagnostics Check		OK	00:21
Air Blank		0.000	00:22
Control Test		0.077	00:22
Air Blank		0.000	00:23
Subject Sample	# 1	0.123	00:23
Breath Volume		1.949	00:23
Air Blank		0.000	00:24
Air Blank		0.000	00:26
Subject Sample	# 2	0.121	00:26
Breath Volume		2.253	00:26
Air Blank		0.000	00:27
Control Test		0.077	00:27
Air Blank		0.000	00:28
Diagnostics Check		OK	00:28

Data Download Date/Time: 09/29/2016 12:28

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PENNINGTON,EMILY,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HOOVER,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 09/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:58

SUBJECT'S LAST NAME: GRACE

DATE/TIME OF TEST: 09/18/2016 01:25:39

Diagnostics Check		OK	01:19
Air Blank		0.000	01:19
Control Test		0.077	01:19
Air Blank		0.000	01:20
Subject Sample	# 1	0.236	01:20
Breath Volume		2.238	01:20
Air Blank		0.000	01:21
Air Blank		0.000	01:23
Subject Sample	# 2	0.220	01:24
Breath Volume		2.855	01:24
Air Blank		0.000	01:24
Control Test		0.076	01:25
Air Blank		0.000	01:25
Diagnostics Check		OK	01:25

Data Download Date/Time: 09/29/2016 12:28

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JOHNSTON,C
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 09/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:33

SUBJECT'S LAST NAME: STAGGERS

DATE/TIME OF TEST: 09/19/2016 01:56:26

Diagnostics Check		OK	01:53	
Air Blank		0.000	01:54	
Control Test		0.077	01:54	
Air Blank		0.000	01:54	
Subject Sample	# 1		01:55	Subject Test Refused
Breath Volume		0.000	01:55	
Air Blank		0.000	01:55	
Control Test		0.077	01:55	
Air Blank		0.000	01:56	
Diagnostics Check		OK	01:56	

Data Download Date/Time: 09/29/2016 12:28

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: RUIZ,R
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 09/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:58

SUBJECT'S LAST NAME: MORENO

DATE/TIME OF TEST: 09/19/2016 04:25:00

Diagnostics Check		OK	04:18
Air Blank		0.000	04:19
Control Test		0.077	04:19
Air Blank		0.000	04:20
Subject Sample	# 1	0.158	04:20
Breath Volume		3.582	04:20
Air Blank		0.000	04:21
Air Blank		0.000	04:23
Subject Sample	# 2	0.142	04:23
Breath Volume		1.824	04:23
Air Blank		0.000	04:24
Control Test		0.076	04:24
Air Blank		0.000	04:24
Diagnostics Check		OK	04:25

Data Download Date/Time: 09/29/2016 12:28

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHITE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 09/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:13

SUBJECT'S LAST NAME: SANCHEZ

DATE/TIME OF TEST: 09/21/2016 02:37:15

Diagnostics Check		OK	02:34	
Air Blank		0.000	02:34	
Control Test		0.077	02:35	
Air Blank		0.000	02:35	
Subject Sample	# 1		02:35	Subject Test Refused
Breath Volume		0.000	02:35	
Air Blank		0.000	02:36	
Control Test		0.077	02:36	
Air Blank		0.000	02:37	
Diagnostics Check		OK	02:37	

Data Download Date/Time: 09/29/2016 12:28

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PENNINGTON,EMILY,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MOORE,A
ARRESTING OFFICER AGENCY: Atlantic Beach PD
LAST AGENCY INSPECTION: 09/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:13

SUBJECT'S LAST NAME: EDDINGS

DATE/TIME OF TEST: 09/23/2016 03:52:03

Diagnostics Check		OK	03:48	
Air Blank		0.000	03:49	
Control Test		0.077	03:49	
Air Blank		0.000	03:49	
Subject Sample	# 1		03:50	Subject Test Refused
Breath Volume		0.000	03:50	
Air Blank		0.000	03:51	
Control Test		0.077	03:51	
Air Blank		0.000	03:51	
Diagnostics Check		OK	03:52	

Data Download Date/Time: 09/29/2016 12:28

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WALLACE,B
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 09/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:06

SUBJECT'S LAST NAME: IMSIROVIC

DATE/TIME OF TEST: 09/24/2016 04:33:10

Diagnostics Check		OK	04:26
Air Blank		0.000	04:27
Control Test		0.077	04:27
Air Blank		0.000	04:27
Subject Sample	# 1	0.060	04:28
Breath Volume		1.589	04:28
Air Blank		0.000	04:29
Air Blank		0.000	04:31
Subject Sample	# 2	0.058	04:31
Breath Volume		1.410	04:31
Air Blank		0.000	04:32
Control Test		0.077	04:32
Air Blank		0.000	04:33
Diagnostics Check		OK	04:33

Data Download Date/Time: 09/29/2016 12:28

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BRYANT,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 09/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:17

SUBJECT'S LAST NAME: VOISIN

DATE/TIME OF TEST: 09/25/2016 01:41:22

Diagnostics Check		OK	01:38	
Air Blank		0.000	01:38	
Control Test		0.076	01:39	
Air Blank		0.000	01:39	
Subject Sample	# 1		01:40	Subject Test Refused
Breath Volume		0.000	01:40	
Air Blank		0.000	01:40	
Control Test		0.078	01:40	
Air Blank		0.000	01:41	
Diagnostics Check		OK	01:41	

Data Download Date/Time: 09/29/2016 12:28

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BROWN,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 09/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: TABOR

DATE/TIME OF TEST: 09/25/2016 03:31:49

Diagnostics Check		OK	03:29	
Air Blank		0.000	03:29	
Control Test		0.077	03:29	
Air Blank		0.000	03:30	
Subject Sample	# 1		03:30	Subject Test Refused
Breath Volume		0.000	03:30	
Air Blank		0.000	03:30	
Control Test		0.078	03:31	
Air Blank		0.000	03:31	
Diagnostics Check		OK	03:31	

Data Download Date/Time: 09/29/2016 12:28

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: KLING,D
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 09/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:11

SUBJECT'S LAST NAME: LITTLEJOHN

DATE/TIME OF TEST: 09/25/2016 23:37:32

Diagnostics Check		OK	23:31
Air Blank		0.000	23:31
Control Test		0.076	23:31
Air Blank		0.000	23:32
Subject Sample	# 1	0.182	23:33
Breath Volume		2.765	23:33
Air Blank		0.000	23:33
Air Blank		0.000	23:35
Subject Sample	# 2	0.183	23:36
Breath Volume		2.769	23:36
Air Blank		0.000	23:36
Control Test		0.077	23:36
Air Blank		0.000	23:37
Diagnostics Check		OK	23:37

Data Download Date/Time: 09/29/2016 12:28

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BROWN,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 09/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: BELL

DATE/TIME OF TEST: 09/26/2016 02:28:37

Diagnostics Check		OK	02:25	
Air Blank		0.000	02:26	
Control Test		0.078	02:26	
Air Blank		0.000	02:27	
Subject Sample	# 1		02:27	Subject Test Refused
Breath Volume		0.000	02:27	
Air Blank		0.000	02:27	
Control Test		0.077	02:28	
Air Blank		0.000	02:28	
Diagnostics Check		OK	02:28	

Data Download Date/Time: 09/29/2016 12:28

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FISETTE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 09/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:57

SUBJECT'S LAST NAME: JOHNS

DATE/TIME OF TEST: 09/28/2016 01:21:02

Diagnostics Check		OK	01:18	
Air Blank		0.000	01:18	
Control Test		0.078	01:18	
Air Blank		0.000	01:19	
Subject Sample	# 1		01:19	Subject Test Refused
Breath Volume		0.000	01:19	
Air Blank		0.000	01:20	
Control Test		0.078	01:20	
Air Blank		0.000	01:20	
Diagnostics Check		OK	01:21	

Data Download Date/Time: 09/29/2016 12:28

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SAUNDERS,H
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 09/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:24

SUBJECT'S LAST NAME: ALDRICH

DATE/TIME OF TEST: 09/28/2016 04:48:47

Diagnostics Check		OK	04:45	
Air Blank		0.000	04:46	
Control Test		0.077	04:46	
Air Blank		0.000	04:47	
Subject Sample	# 1		04:47	Subject Test Refused
Breath Volume		0.000	04:47	
Air Blank		0.000	04:47	
Control Test		0.078	04:48	
Air Blank		0.000	04:48	
Diagnostics Check		OK	04:48	

Data Download Date/Time: 09/29/2016 12:28

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIALKOSKI,WILLIAM,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: GARCIA,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 09/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 18:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:01

SUBJECT'S LAST NAME: MONTELEONE

DATE/TIME OF TEST: 09/28/2016 19:31:54

Diagnostics Check		OK	19:21
Air Blank		0.000	19:22
Control Test		0.077	19:22
Air Blank		0.000	19:22
Subject Sample	# 1	0.093	19:24
Breath Volume		1.835	19:24
Air Blank		0.000	19:24
Air Blank		0.000	19:26
Subject Sample	# 2	0.069	19:27
Breath Volume		1.300	19:27
Air Blank		0.000	19:27
Air Blank		0.000	19:29
Subject Sample	# 3	0.080	19:30
Breath Volume		2.050	19:30
Air Blank		0.000	19:30
Control Test		0.077	19:31
Air Blank		0.000	19:31
Diagnostics Check		OK	19:31

Data Download Date/Time: 09/29/2016 12:28

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CAMACHO,K
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 09/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:08

SUBJECT'S LAST NAME: FOUNTAIN

DATE/TIME OF TEST: 09/29/2016 01:34:53

Diagnostics Check		OK	01:28
Air Blank		0.000	01:28
Control Test		0.077	01:29
Air Blank		0.000	01:29
Subject Sample	# 1	0.125	01:30
Breath Volume		1.257	01:30
Air Blank		0.000	01:30
Air Blank		0.000	01:32
Subject Sample	# 2	0.128	01:33
Breath Volume		1.203	01:33
Air Blank		0.000	01:33
Control Test		0.076	01:34
Air Blank		0.000	01:34
Diagnostics Check		OK	01:34

Data Download Date/Time: 09/29/2016 12:28

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHITE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 09/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:38

SUBJECT'S LAST NAME: CHAMBERS

DATE/TIME OF TEST: 09/29/2016 03:01:32

Diagnostics Check		OK	02:58	
Air Blank		0.000	02:59	
Control Test		0.077	02:59	
Air Blank		0.000	02:59	
Subject Sample	# 1		03:00	Subject Test Refused
Breath Volume		0.000	03:00	
Air Blank		0.000	03:00	
Control Test		0.077	03:00	
Air Blank		0.000	03:01	
Diagnostics Check		OK	03:01	

Data Download Date/Time: 09/29/2016 12:28

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BRETT,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 09/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:23

SUBJECT'S LAST NAME: HORR

DATE/TIME OF TEST: 09/29/2016 04:50:17

Diagnostics Check		OK	04:43
Air Blank		0.000	04:43
Control Test		0.077	04:44
Air Blank		0.000	04:44
Subject Sample	# 1	0.153	04:45
Breath Volume		1.480	04:45
Air Blank		0.000	04:46
Air Blank		0.000	04:48
Subject Sample	# 2	0.144	04:48
Breath Volume		1.558	04:48
Air Blank		0.000	04:49
Control Test		0.077	04:49
Air Blank		0.000	04:50
Diagnostics Check		OK	04:50

Data Download Date/Time: 09/29/2016 12:28

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: STANLEY,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 09/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:22

SUBJECT'S LAST NAME: STEELE

DATE/TIME OF TEST: 09/30/2016 01:49:09

Diagnostics Check		OK	01:42
Air Blank		0.000	01:43
Control Test		0.077	01:43
Air Blank		0.000	01:44
Subject Sample	# 1	0.137	01:44
Breath Volume		3.539	01:44
Air Blank		0.000	01:45
Air Blank		0.000	01:47
Subject Sample	# 2	0.138	01:47
Breath Volume		2.851	01:47
Air Blank		0.000	01:48
Control Test		0.077	01:48
Air Blank		0.000	01:49
Diagnostics Check		OK	01:49

Data Download Date/Time: 10/14/2016 09:59

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBER,JOSEPH,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHITE,J.M.
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 09/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:37

SUBJECT'S LAST NAME: MALLS

DATE/TIME OF TEST: 09/30/2016 05:04:29

Diagnostics Check		OK	04:58
Air Blank		0.000	04:58
Control Test		0.077	04:58
Air Blank		0.000	04:59
Subject Sample	# 1	0.196	04:59
Breath Volume		1.222	04:59
Air Blank		0.000	05:00
Air Blank		0.000	05:02
Subject Sample	# 2	0.195	05:02
Breath Volume		1.949	05:02
Air Blank		0.000	05:03
Control Test		0.077	05:03
Air Blank		0.000	05:04
Diagnostics Check		OK	05:04

Data Download Date/Time: 10/14/2016 09:59

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BRETT,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 09/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:17

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 09/30/2016 05:44:13

Diagnostics Check		OK	05:38
Air Blank		0.000	05:38
Control Test		0.077	05:38
Air Blank		0.000	05:39
Subject Sample	# 1	0.149	05:39
Breath Volume		2.261	05:39
Air Blank		0.000	05:40
Air Blank		0.000	05:42
Subject Sample	# 2	0.154	05:42
Breath Volume		2.929	05:42
Air Blank		0.000	05:43
Control Test		0.077	05:43
Air Blank		0.000	05:44
Diagnostics Check		OK	05:44

Data Download Date/Time: 10/14/2016 09:59

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001109

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,ALAIN,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: HERNANDEZ,ALAIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 08/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 16014080A2
DRY GAS EXPIRY DATE: 07/05/2016
VIOLATION CODE: DUI
ARREST TIME: 00:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:53

SUBJECT'S LAST NAME: PFEIFFER

DATE/TIME OF TEST: 09/24/2016 00:37:30

Diagnostics Check		OK	00:28
Air Blank		0.000	00:29
Control Test		0.077	00:29
Air Blank		0.000	00:30
Subject Sample	# 1	0.091	00:32
Breath Volume		1.144	00:32
Air Blank		0.000	00:33
Air Blank		0.000	00:35
Subject Sample	# 2	0.089	00:35
Breath Volume		1.300	00:35
Air Blank		0.000	00:36
Control Test		0.076	00:36
Air Blank		0.000	00:37
Diagnostics Check		OK	00:37

Data Download Date/Time: 09/28/2016 22:36

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001110

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADDISON,RAYMOND,E
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: MARRERO,JUAN
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 08/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1878845
DRY GAS EXPIRY DATE: 06/22/2018
VIOLATION CODE: DUI
ARREST TIME: 23:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:36

SUBJECT'S LAST NAME: ROSE

DATE/TIME OF TEST: 09/04/2016 00:03:17

Diagnostics Check		OK	23:56
Air Blank		0.000	23:56
Control Test		0.078	23:57
Air Blank		0.000	23:57
Subject Sample	# 2	0.163	23:58
Breath Volume		3.464	23:58
Air Blank		0.000	23:59
Air Blank		0.000	00:01
Subject Sample	# 1	0.160	00:01
Breath Volume		2.714	00:01
Air Blank		0.000	00:02
Control Test		0.079	00:02
Air Blank		0.000	00:03
Diagnostics Check		OK	00:03

Data Download Date/Time: 09/16/2016 11:37

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001110

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADDISON,RAYMOND,E
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: GARCIA,JOHAN
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 08/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1878845
DRY GAS EXPIRY DATE: 06/22/2018
VIOLATION CODE: DUI
ARREST TIME: 05:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:06

SUBJECT'S LAST NAME: VAN KANTEN

DATE/TIME OF TEST: 09/05/2016 05:32:20

Diagnostics Check		OK	05:28	
Air Blank		0.000	05:29	
Control Test		0.080	05:29	
Air Blank		0.000	05:30	
Subject Sample	# 1		05:30	Subject Test Refused
Breath Volume		0.000	05:30	
Air Blank		0.000	05:31	
Control Test		0.078	05:31	
Air Blank		0.000	05:32	
Diagnostics Check		OK	05:32	

Data Download Date/Time: 09/16/2016 11:37

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001111

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILEY,DAVID,W
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: WILEY,DAVID
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 08/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:18

SUBJECT'S LAST NAME: SCURLOCK

DATE/TIME OF TEST: 09/06/2016 02:41:26

Diagnostics Check		OK	02:38	
Air Blank		0.000	02:39	
Control Test		0.080	02:39	
Air Blank		0.000	02:39	
Subject Sample	# 1		02:40	Subject Test Refused
Breath Volume		0.000	02:40	
Air Blank		0.000	02:40	
Control Test		0.079	02:40	
Air Blank		0.000	02:41	
Diagnostics Check		OK	02:41	

Data Download Date/Time: 09/27/2016 20:52

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001111

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAILEY,RAYMOND,F
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BAILEY,RAYMOND
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 08/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: MILLS

DATE/TIME OF TEST: 09/11/2016 02:53:34

Diagnostics Check		OK	02:50	
Air Blank		0.000	02:51	
Control Test		0.079	02:51	
Air Blank		0.000	02:51	
Subject Sample	# 1		02:52	Subject Test Refused
Breath Volume		0.000	02:52	
Air Blank		0.000	02:52	
Control Test		0.080	02:53	
Air Blank		0.000	02:53	
Diagnostics Check		OK	02:53	

Data Download Date/Time: 09/27/2016 20:52

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001111

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAILEY,RAYMOND,F
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BAILEY,RAYMOND
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 08/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 09:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:25

SUBJECT'S LAST NAME: ALEXANDER

DATE/TIME OF TEST: 09/13/2016 10:57:53

Diagnostics Check		OK	10:45	
Air Blank		0.000	10:46	
Control Test		0.078	10:46	
Air Blank		0.000	10:47	
Subject Sample	# 1	0.110	10:50	Volume Not Met
Breath Volume		0.035	10:50	
Air Blank		0.000	10:50	
Air Blank		0.000	10:52	
Subject Sample	# 2	0.121	10:53	
Breath Volume		1.765	10:53	
Air Blank		0.000	10:53	
Air Blank		0.000	10:55	
Subject Sample	# 3	0.116	10:56	
Breath Volume		1.750	10:56	
Air Blank		0.000	10:56	
Control Test		0.079	10:57	
Air Blank		0.000	10:57	
Diagnostics Check		OK	10:57	

Data Download Date/Time: 09/27/2016 20:52

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001111

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VENNELL,JOHN,B
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: VENNELL,JOHN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 08/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: HARMOM

DATE/TIME OF TEST: 09/21/2016 00:56:22

Diagnostics Check		OK	00:44	
Air Blank		0.000	00:45	
Control Test		0.079	00:45	
Air Blank		0.000	00:45	
Subject Sample	# 1		00:49	No Sample Provided
Breath Volume		0.000	00:49	
Air Blank		0.000	00:49	
Air Blank		0.000	00:51	
Subject Sample	# 2	0.163	00:51	
Breath Volume		1.480	00:51	
Air Blank		0.000	00:52	
Air Blank		0.000	00:54	
Subject Sample	# 3	0.176	00:54	
Breath Volume		2.753	00:54	
Air Blank		0.000	00:55	
Control Test		0.079	00:55	
Air Blank		0.000	00:56	
Diagnostics Check		OK	00:56	

Data Download Date/Time: 09/27/2016 20:52

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001113

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,ALAIN,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: HERNANDEZ,ALAIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 08/02/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 477627
DRY GAS EXPIRY DATE: 06/29/2019
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:36

SUBJECT'S LAST NAME: VENTURA-GONZALEZ

DATE/TIME OF TEST: 09/03/2016 03:02:49

Diagnostics Check		OK	02:56
Air Blank		0.000	02:57
Control Test		0.080	02:57
Air Blank		0.000	02:58
Subject Sample	# 1	0.178	02:58
Breath Volume		1.953	02:58
Air Blank		0.000	02:59
Air Blank		0.000	03:00
Subject Sample	# 2	0.171	03:01
Breath Volume		1.691	03:01
Air Blank		0.000	03:01
Control Test		0.079	03:02
Air Blank		0.000	03:02
Diagnostics Check		OK	03:02

Data Download Date/Time: 09/28/2016 22:32

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001113

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,ALAIN,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: HERNANDEZ,ALAIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 08/02/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 477627
DRY GAS EXPIRY DATE: 06/29/2019
VIOLATION CODE: DUI
ARREST TIME: 03:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:21

SUBJECT'S LAST NAME: DELGADO-ROSALES

DATE/TIME OF TEST: 09/05/2016 03:47:31

Diagnostics Check		OK	03:41
Air Blank		0.000	03:42
Control Test		0.080	03:42
Air Blank		0.000	03:42
Subject Sample	# 1	0.140	03:43
Breath Volume		1.335	03:43
Air Blank		0.000	03:43
Air Blank		0.000	03:45
Subject Sample	# 2	0.134	03:46
Breath Volume		1.324	03:46
Air Blank		0.000	03:46
Control Test		0.080	03:46
Air Blank		0.000	03:47
Diagnostics Check		OK	03:47

Data Download Date/Time: 09/28/2016 22:32

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001113

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,ALAIN,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: HERNANDEZ,ALAIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 08/02/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 477627
DRY GAS EXPIRY DATE: 06/29/2019
VIOLATION CODE: DUI
ARREST TIME: 16:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:31

SUBJECT'S LAST NAME: GRUBBS

DATE/TIME OF TEST: 09/05/2016 16:58:33

Diagnostics Check		OK	16:52
Air Blank		0.000	16:53
Control Test		0.080	16:53
Air Blank		0.000	16:53
Subject Sample	# 1	0.000	16:54
Breath Volume		2.105	16:54
Air Blank		0.000	16:54
Air Blank		0.000	16:56
Subject Sample	# 2	0.000	16:57
Breath Volume		2.070	16:57
Air Blank		0.000	16:57
Control Test		0.079	16:57
Air Blank		0.000	16:58
Diagnostics Check		OK	16:58

Data Download Date/Time: 09/28/2016 22:32

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001113

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,ALAIN,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: HERNANDEZ,ALAIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 08/02/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 477627
DRY GAS EXPIRY DATE: 06/29/2019
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:43

SUBJECT'S LAST NAME: HERNANDEZ-TORRES

DATE/TIME OF TEST: 09/11/2016 02:09:15

Diagnostics Check		OK	02:03
Air Blank		0.000	02:03
Control Test		0.080	02:03
Air Blank		0.000	02:04
Subject Sample	# 1	0.110	02:05
Breath Volume		2.105	02:05
Air Blank		0.000	02:05
Air Blank		0.000	02:07
Subject Sample	# 2	0.103	02:07
Breath Volume		2.546	02:07
Air Blank		0.000	02:08
Control Test		0.080	02:08
Air Blank		0.000	02:09
Diagnostics Check		OK	02:09

Data Download Date/Time: 09/28/2016 22:32

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001113

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,ALAIN,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: HERNANDEZ,ALAIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 08/02/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 477627
DRY GAS EXPIRY DATE: 06/29/2019
VIOLATION CODE: DUI
ARREST TIME: 01:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:22

SUBJECT'S LAST NAME: GASPAR-GARCIA

DATE/TIME OF TEST: 09/18/2016 01:56:31

Diagnostics Check		OK	01:50
Air Blank		0.000	01:50
Control Test		0.080	01:51
Air Blank		0.000	01:51
Subject Sample	# 1	0.170	01:52
Breath Volume		1.281	01:52
Air Blank		0.000	01:52
Air Blank		0.000	01:54
Subject Sample	# 2	0.167	01:55
Breath Volume		1.187	01:55
Air Blank		0.000	01:55
Control Test		0.080	01:55
Air Blank		0.000	01:56
Diagnostics Check		OK	01:56

Data Download Date/Time: 09/28/2016 22:32

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001113

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,ALAIN,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: HERNANDEZ,ALAIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 08/02/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 477627
DRY GAS EXPIRY DATE: 06/29/2019
VIOLATION CODE: DUI
ARREST TIME: 02:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:36

SUBJECT'S LAST NAME: HOUSTON

DATE/TIME OF TEST: 09/20/2016 03:04:19

Diagnostics Check		OK	02:58
Air Blank		0.000	02:58
Control Test		0.081	02:58
Air Blank		0.000	02:59
Subject Sample	# 1	0.195	02:59
Breath Volume		1.375	02:59
Air Blank		0.000	03:00
Air Blank		0.000	03:02
Subject Sample	# 2	0.199	03:02
Breath Volume		1.902	03:02
Air Blank		0.000	03:03
Control Test		0.081	03:03
Air Blank		0.000	03:04
Diagnostics Check		OK	03:04

Data Download Date/Time: 09/28/2016 22:32

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001113

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,ALAIN,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SNAGGS,SHELDON
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 08/02/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 477627
DRY GAS EXPIRY DATE: 06/29/2019
VIOLATION CODE: DUI
ARREST TIME: 00:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: OLSON

DATE/TIME OF TEST: 09/22/2016 01:29:54

Diagnostics Check		OK	01:15	
Air Blank		0.000	01:16	
Control Test		0.080	01:16	
Air Blank		0.000	01:17	
Subject Sample	# 1	0.144	01:20	Volume Not Met
Breath Volume		0.121	01:20	
Air Blank		0.000	01:20	
Air Blank		0.000	01:22	
Subject Sample	# 2	0.143	01:25	
Breath Volume		1.769	01:25	
Air Blank		0.000	01:25	
Air Blank		0.000	01:27	
Subject Sample	# 3	0.146	01:28	
Breath Volume		1.269	01:28	
Air Blank		0.000	01:28	
Control Test		0.080	01:29	
Air Blank		0.000	01:29	
Diagnostics Check		OK	01:29	

Data Download Date/Time: 09/28/2016 22:32

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001113

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,ALAIN,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: HERNANDEZ,ALAIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 08/02/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 477627
DRY GAS EXPIRY DATE: 06/29/2019
VIOLATION CODE: DUI
ARREST TIME: 00:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:53

SUBJECT'S LAST NAME: PFEIFFER

DATE/TIME OF TEST: 09/24/2016 00:20:00

Diagnostics Check	OK	00:18	
Air Blank		00:19	Ambient Fail
Air Blank		00:20	Ambient Fail

Data Download Date/Time: 09/28/2016 22:32

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001113

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,ALAIN,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: HERNANDEZ,ALAIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 08/02/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 477627
DRY GAS EXPIRY DATE: 06/29/2019
VIOLATION CODE: DUI
ARREST TIME: 00:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:53

SUBJECT'S LAST NAME: PFEIFFER

DATE/TIME OF TEST: 09/24/2016 00:22:31

Diagnostics Check	OK	00:21	
Air Blank		00:22	Ambient Fail
Air Blank	0.000	00:22	

Data Download Date/Time: 09/28/2016 22:32

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001113

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,ALAIN,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: HERNANDEZ,ALAIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 08/02/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 477627
DRY GAS EXPIRY DATE: 06/29/2019
VIOLATION CODE: DUI
ARREST TIME: 03:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: LILLEY

DATE/TIME OF TEST: 09/24/2016 03:19:40

Diagnostics Check		OK	03:12
Air Blank		0.000	03:13
Control Test		0.080	03:13
Air Blank		0.000	03:14
Subject Sample	# 1	0.177	03:14
Breath Volume		1.554	03:14
Air Blank		0.000	03:15
Air Blank		0.000	03:17
Subject Sample	# 2	0.169	03:17
Breath Volume		1.718	03:17
Air Blank		0.000	03:18
Control Test		0.081	03:19
Air Blank		0.000	03:19
Diagnostics Check		OK	03:19

Data Download Date/Time: 09/28/2016 22:32

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001113

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,ALAIN,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: DOKENDORF,BRYAN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 08/02/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 477627
DRY GAS EXPIRY DATE: 06/29/2019
VIOLATION CODE: DUI
ARREST TIME: 23:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: BUSTER JR

DATE/TIME OF TEST: 09/24/2016 23:39:37

Diagnostics Check		OK	23:33
Air Blank		0.000	23:34
Control Test		0.081	23:34
Air Blank		0.000	23:35
Subject Sample	# 1	0.000	23:35
Breath Volume		2.792	23:35
Air Blank		0.000	23:35
Air Blank		0.000	23:37
Subject Sample	# 2	0.000	23:38
Breath Volume		2.531	23:38
Air Blank		0.000	23:38
Control Test		0.081	23:39
Air Blank		0.000	23:39
Diagnostics Check		OK	23:39

Data Download Date/Time: 09/28/2016 22:32

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001113

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,ALAIN,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: HERNANDEZ,ALAIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 08/02/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 477627
DRY GAS EXPIRY DATE: 06/29/2019
VIOLATION CODE: DUI
ARREST TIME: 01:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:22

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 09/25/2016 01:57:05

Diagnostics Check		OK	01:51
Air Blank		0.000	01:51
Control Test		0.080	01:51
Air Blank		0.000	01:52
Subject Sample	# 1	0.132	01:52
Breath Volume		2.410	01:52
Air Blank		0.000	01:53
Air Blank		0.000	01:55
Subject Sample	# 2	0.116	01:55
Breath Volume		2.265	01:55
Air Blank		0.000	01:56
Control Test		0.080	01:56
Air Blank		0.000	01:56
Diagnostics Check		OK	01:57

Data Download Date/Time: 09/28/2016 22:32

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001113

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,ALAIN,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: HERNANDEZ,ALAIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 09/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 477627
DRY GAS EXPIRY DATE: 06/29/2019
VIOLATION CODE: DUI
ARREST TIME: 00:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:53

SUBJECT'S LAST NAME: GREEN

DATE/TIME OF TEST: 09/29/2016 01:19:02

Diagnostics Check		OK	01:13	
Air Blank		0.000	01:13	
Control Test		0.081	01:13	
Air Blank		0.000	01:14	
Subject Sample	# 1	0.182	01:14	
Breath Volume		1.898	01:14	
Air Blank		0.000	01:15	
Air Blank		0.000	01:17	
Subject Sample	# 2	0.176	01:17	
Breath Volume		1.308	01:17	
Air Blank		0.000	01:18	
Control Test		0.081	01:18	
Air Blank			01:18	RFI Detect
Air Blank		0.000	01:19	

Data Download Date/Time: 10/26/2016 22:02

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001113

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,ALAIN,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: HERNANDEZ,ALAIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 09/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 477627
DRY GAS EXPIRY DATE: 06/29/2019
VIOLATION CODE: DUI
ARREST TIME: 00:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:53

SUBJECT'S LAST NAME: GREEN

DATE/TIME OF TEST: 09/29/2016 01:26:50

Diagnostics Check		OK	01:20
Air Blank		0.000	01:21
Control Test		0.080	01:21
Air Blank		0.000	01:21
Subject Sample	# 1	0.172	01:22
Breath Volume		1.492	01:22
Air Blank		0.000	01:23
Air Blank		0.000	01:25
Subject Sample	# 2	0.175	01:25
Breath Volume		1.699	01:25
Air Blank		0.000	01:25
Control Test		0.080	01:26
Air Blank		0.000	01:26
Diagnostics Check		OK	01:26

Data Download Date/Time: 10/26/2016 22:02

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001121

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORTES,HECTOR,E
BREATH TEST OPERATOR AGENCY: FHP Troop K
ARRESTING OFFICER: CORTES,HECTOR
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 08/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: SANCHEZ

DATE/TIME OF TEST: 09/11/2016 03:40:09

Diagnostics Check		OK	03:34	
Air Blank		0.000	03:35	
Control Test		0.079	03:35	
Air Blank		0.000	03:36	
Subject Sample	# 1		03:38	Subject Test Refused
Breath Volume		0.000	03:38	
Air Blank		0.000	03:39	
Control Test		0.080	03:39	
Air Blank		0.000	03:40	
Diagnostics Check		OK	03:40	

Data Download Date/Time: 09/16/2016 08:59

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001123

INSTRUMENT REGISTERED TO: FHP TROOP D ORLANDO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOONDA,SHUAIB,
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: MOONDA,SHUAIB
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 08/31/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20314080A5
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 11:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: STEVENS

DATE/TIME OF TEST: 09/05/2016 00:53:07

Diagnostics Check		OK	00:44
Air Blank		0.000	00:45
Control Test		0.078	00:45
Air Blank		0.000	00:45
Subject Sample	# 1	0.176	00:47
Breath Volume		2.531	00:47
Air Blank		0.000	00:48
Air Blank		0.000	00:50
Subject Sample	# 2	0.182	00:51
Breath Volume		3.773	00:51
Air Blank		0.000	00:52
Control Test		0.078	00:52
Air Blank		0.000	00:53
Diagnostics Check		OK	00:53

Data Download Date/Time: 10/01/2016 00:10

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001123

INSTRUMENT REGISTERED TO: FHP TROOP D ORLANDO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOONDA,SHUAIB,
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: MOONDA,SHUAIB
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 08/31/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20314080A5
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 05:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:32

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 09/05/2016 06:01:18

Diagnostics Check		OK	05:54
Air Blank		0.000	05:55
Control Test		0.079	05:55
Air Blank		0.000	05:56
Subject Sample	# 1	0.285	05:56
Breath Volume		2.402	05:56
Air Blank		0.000	05:57
Air Blank		0.000	05:59
Subject Sample	# 2	0.287	05:59
Breath Volume		3.117	05:59
Air Blank		0.000	06:00
Control Test		0.078	06:00
Air Blank		0.000	06:01
Diagnostics Check		OK	06:01

Data Download Date/Time: 10/01/2016 00:10

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001128

INSTRUMENT REGISTERED TO: PUTNAM COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEWIS,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Putnam County SO
ARRESTING OFFICER: LEWIS,JUSTIN
ARRESTING OFFICER AGENCY: Putnam County SO
LAST AGENCY INSPECTION: 08/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07515080A4
DRY GAS EXPIRY DATE: 04/05/2017
VIOLATION CODE: DUI
ARREST TIME: 12:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:43

SUBJECT'S LAST NAME: KLINE

DATE/TIME OF TEST: 09/10/2016 01:19:14

Diagnostics Check		OK	01:12
Air Blank		0.000	01:12
Control Test		0.077	01:12
Air Blank		0.000	01:13
Subject Sample	# 1	0.137	01:14
Breath Volume		3.128	01:14
Air Blank		0.000	01:15
Air Blank		0.000	01:17
Subject Sample	# 2	0.131	01:17
Breath Volume		1.246	01:17
Air Blank		0.000	01:18
Control Test		0.077	01:18
Air Blank		0.000	01:19
Diagnostics Check		OK	01:19

Data Download Date/Time: 09/29/2016 20:38

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001128

INSTRUMENT REGISTERED TO: PUTNAM COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,TIFFANY,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: WILLIAMS,TIFFANY
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 08/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07515080A4
DRY GAS EXPIRY DATE: 04/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: TENNEY

DATE/TIME OF TEST: 09/14/2016 22:54:21

Diagnostics Check		OK	22:41	
Air Blank		0.000	22:41	
Control Test		0.077	22:41	
Air Blank		0.000	22:42	
Subject Sample	# 1	0.105	22:45	Volume Not Met
Breath Volume		0.945	22:45	
Air Blank		0.000	22:46	
Air Blank		0.000	22:48	
Subject Sample	# 2	0.116	22:48	
Breath Volume		1.250	22:48	
Air Blank		0.000	22:49	
Air Blank		0.000	22:51	
Subject Sample	# 3	0.111	22:52	
Breath Volume		1.222	22:52	
Air Blank		0.000	22:53	
Control Test		0.076	22:53	
Air Blank		0.000	22:54	
Diagnostics Check		OK	22:54	

Data Download Date/Time: 09/29/2016 20:38

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001128

INSTRUMENT REGISTERED TO: PUTNAM COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEWIS,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Putnam County SO
ARRESTING OFFICER: VALDIVIA,SHAUN
ARRESTING OFFICER AGENCY: Putnam County SO
LAST AGENCY INSPECTION: 08/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07515080A4
DRY GAS EXPIRY DATE: 04/05/2017
VIOLATION CODE: DUI
ARREST TIME: 09:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: BAER

DATE/TIME OF TEST: 09/15/2016 23:13:00

Diagnostics Check		OK	23:04	
Air Blank		0.000	23:04	
Control Test		0.077	23:05	
Air Blank		0.000	23:05	
Subject Sample	# 1	0.252	23:07	
Breath Volume		1.160	23:07	
Air Blank		0.000	23:08	
Air Blank		0.000	23:09	
Subject Sample	# 2	0.239	23:11	
Breath Volume		1.148	23:11	
Air Blank		0.000	23:12	
Control Test		0.074	23:12	Control Outside Tolerance
Air Blank		0.000	23:13	

Data Download Date/Time: 09/29/2016 20:38

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001128

INSTRUMENT REGISTERED TO: PUTNAM COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VEGA,KELVIN,
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: VEGA,KELVIN
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 08/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07515080A4
DRY GAS EXPIRY DATE: 04/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: BOYER

DATE/TIME OF TEST: 09/23/2016 02:21:36

Diagnostics Check		OK	02:14
Air Blank		0.000	02:14
Control Test		0.077	02:14
Air Blank		0.000	02:15
Subject Sample	# 1	0.127	02:16
Breath Volume		1.378	02:16
Air Blank		0.000	02:16
Air Blank		0.000	02:18
Subject Sample	# 2	0.128	02:20
Breath Volume		3.101	02:20
Air Blank		0.000	02:20
Control Test		0.076	02:21
Air Blank		0.000	02:21
Diagnostics Check		OK	02:21

Data Download Date/Time: 09/29/2016 20:38

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001129

INSTRUMENT REGISTERED TO: PUTNAM COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BEDENBAUGH,JUSTIN,H
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: BEDENBAUGH,JUSTIN
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 08/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07515080A4
DRY GAS EXPIRY DATE: 04/05/2017
VIOLATION CODE: DUI
ARREST TIME: 18:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:46

SUBJECT'S LAST NAME: HAMMITT

DATE/TIME OF TEST: 09/13/2016 19:29:01

Diagnostics Check		OK	19:17	
Air Blank		0.000	19:18	
Control Test		0.078	19:18	
Air Blank		0.000	19:18	
Subject Sample	# 1		19:22	No Sample Provided
Breath Volume		0.000	19:22	
Air Blank		0.000	19:22	
Air Blank		0.000	19:24	
Subject Sample	# 2		19:27	No Sample Provided
Breath Volume		0.000	19:27	
Air Blank		0.000	19:28	
Control Test		0.081	19:28	
Air Blank		0.000	19:28	
Diagnostics Check		OK	19:29	

Data Download Date/Time: 09/29/2016 20:23

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001130

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANSARD,JOSHUA,D
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: MYERS,MATTHEW
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 08/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 20:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:30

SUBJECT'S LAST NAME: DIAZ-MARQUEZ

DATE/TIME OF TEST: 09/01/2016 21:02:57

Diagnostics Check		OK	20:56
Air Blank		0.000	20:57
Control Test		0.079	20:57
Air Blank		0.000	20:58
Subject Sample	# 1	0.213	20:58
Breath Volume		2.222	20:58
Air Blank		0.000	20:59
Air Blank		0.000	21:01
Subject Sample	# 2	0.205	21:01
Breath Volume		1.523	21:01
Air Blank		0.000	21:02
Control Test		0.079	21:02
Air Blank		0.000	21:02
Diagnostics Check		OK	21:02

Data Download Date/Time: 09/15/2016 12:46

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001130

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORBIN,STEPHEN,A
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: MYERS,MATTHEW
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 08/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 20:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:15

SUBJECT'S LAST NAME: TARAS

DATE/TIME OF TEST: 09/02/2016 20:50:23

Diagnostics Check		OK	20:41
Air Blank		0.000	20:41
Control Test		0.080	20:41
Air Blank		0.000	20:42
Subject Sample	# 1	0.241	20:44
Breath Volume		1.386	20:44
Air Blank		0.000	20:44
Air Blank		0.000	20:46
Subject Sample	# 2	0.243	20:48
Breath Volume		1.570	20:48
Air Blank		0.000	20:49
Control Test		0.079	20:49
Air Blank		0.000	20:50
Diagnostics Check		OK	20:50

Data Download Date/Time: 09/15/2016 12:46

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001130

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORBIN,STEPHEN,A
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: CORBIN,STEPHEN
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 08/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 23:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 09/03/2016 00:43:10

Diagnostics Check		OK	00:36
Air Blank		0.000	00:37
Control Test		0.080	00:37
Air Blank		0.000	00:38
Subject Sample	# 1	0.232	00:38
Breath Volume		2.035	00:38
Air Blank		0.000	00:39
Air Blank		0.000	00:41
Subject Sample	# 2	0.225	00:41
Breath Volume		1.324	00:41
Air Blank		0.000	00:42
Control Test		0.079	00:42
Air Blank		0.000	00:43
Diagnostics Check		OK	00:43

Data Download Date/Time: 09/15/2016 12:46

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001130

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,MATTHEW,A
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: MYERS,MATTHEW
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 08/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:43

SUBJECT'S LAST NAME: JACQUES

DATE/TIME OF TEST: 09/03/2016 02:16:11

Diagnostics Check		OK	02:09
Air Blank		0.000	02:10
Control Test		0.080	02:10
Air Blank		0.000	02:10
Subject Sample	# 1	0.112	02:11
Breath Volume		1.726	02:11
Air Blank		0.000	02:12
Air Blank		0.000	02:14
Subject Sample	# 2	0.113	02:14
Breath Volume		2.582	02:14
Air Blank		0.000	02:15
Control Test		0.080	02:15
Air Blank		0.000	02:16
Diagnostics Check		OK	02:16

Data Download Date/Time: 09/15/2016 12:46

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001130

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,MATTHEW,A
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: MYERS,MATTHEW
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 08/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 00:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: MCQUEEN

DATE/TIME OF TEST: 09/04/2016 01:57:47

Diagnostics Check		OK	01:46
Air Blank		0.000	01:47
Control Test		0.079	01:47
Air Blank		0.000	01:48
Subject Sample	# 1	0.174	01:48
Breath Volume		1.957	01:48
Air Blank		0.000	01:49
Air Blank		0.000	01:51
Subject Sample	# 2	0.148	01:52
Breath Volume		1.238	01:52
Air Blank		0.000	01:53
Air Blank		0.000	01:55
Subject Sample	# 3	0.175	01:56
Breath Volume		1.527	01:56
Air Blank		0.000	01:56
Control Test		0.079	01:57
Air Blank		0.000	01:57
Diagnostics Check		OK	01:57

Data Download Date/Time: 09/15/2016 12:46

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001130

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,MATTHEW,A
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: MYERS,MATTHEW
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 08/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 01:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:46

SUBJECT'S LAST NAME: TEMPESTA

DATE/TIME OF TEST: 09/15/2016 03:13:38

Diagnostics Check		OK	03:06
Air Blank		0.000	03:06
Control Test		0.079	03:07
Air Blank		0.000	03:07
Subject Sample	# 1	0.119	03:08
Breath Volume		3.210	03:08
Air Blank		0.000	03:09
Air Blank		0.000	03:11
Subject Sample	# 2	0.122	03:12
Breath Volume		1.425	03:12
Air Blank		0.000	03:12
Control Test		0.079	03:13
Air Blank		0.000	03:13
Diagnostics Check		OK	03:13

Data Download Date/Time: 09/15/2016 12:46

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001130

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PECK,KEITH,D
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: PECK,KEITH
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 09/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: EWART

DATE/TIME OF TEST: 09/19/2016 01:42:43

Diagnostics Check		OK	01:36
Air Blank		0.000	01:36
Control Test		0.079	01:36
Air Blank		0.000	01:37
Subject Sample	# 1	0.155	01:38
Breath Volume		3.296	01:38
Air Blank		0.000	01:38
Air Blank		0.000	01:40
Subject Sample	# 2	0.152	01:41
Breath Volume		2.410	01:41
Air Blank		0.000	01:41
Control Test		0.079	01:42
Air Blank		0.000	01:42
Diagnostics Check		OK	01:42

Data Download Date/Time: 10/19/2016 19:30

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001131

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FROST,JOSEPH,W
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: BRYAN,STEVEN
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 08/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 22:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:31

SUBJECT'S LAST NAME: CURTIS

DATE/TIME OF TEST: 09/03/2016 00:04:51

Diagnostics Check		OK	23:57
Air Blank		0.000	23:57
Control Test		0.079	23:57
Air Blank		0.000	23:58
Subject Sample	# 1	0.206	00:00
Breath Volume		1.640	00:00
Air Blank		0.000	00:00
Air Blank		0.000	00:02
Subject Sample	# 2	0.224	00:03
Breath Volume		1.750	00:03
Air Blank		0.000	00:03
Control Test		0.080	00:04
Air Blank		0.000	00:04
Diagnostics Check		OK	00:04

Data Download Date/Time: 09/15/2016 09:25

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001131

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRYAN,STEVEN,A
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: STONE,JIMMIE
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 08/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 02:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:52

SUBJECT'S LAST NAME: GOIRIENA

DATE/TIME OF TEST: 09/04/2016 03:19:57

Diagnostics Check		OK	03:12
Air Blank		0.000	03:12
Control Test		0.080	03:13
Air Blank		0.000	03:13
Subject Sample	# 1	0.186	03:15
Breath Volume		1.238	03:15
Air Blank		0.000	03:15
Air Blank		0.000	03:17
Subject Sample	# 2	0.188	03:18
Breath Volume		1.593	03:18
Air Blank		0.000	03:19
Control Test		0.080	03:19
Air Blank		0.000	03:19
Diagnostics Check		OK	03:19

Data Download Date/Time: 09/15/2016 09:25

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001131

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CALKINS,BRIAN,M
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: GODENZI,CAMERON
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 08/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 23:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:48

SUBJECT'S LAST NAME: HARGROVE

DATE/TIME OF TEST: 09/12/2016 00:23:56

Diagnostics Check		OK	00:17
Air Blank		0.000	00:18
Control Test		0.080	00:18
Air Blank		0.000	00:19
Subject Sample	# 1	0.112	00:19
Breath Volume		2.406	00:19
Air Blank		0.000	00:20
Air Blank		0.000	00:21
Subject Sample	# 2	0.108	00:22
Breath Volume		1.683	00:22
Air Blank		0.000	00:23
Control Test		0.080	00:23
Air Blank		0.000	00:23
Diagnostics Check		OK	00:23

Data Download Date/Time: 09/15/2016 09:25

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.