

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: POLIZZI,CHRISTINE,A  
BREATH TEST OPERATOR AGENCY: Martin County SO  
ARRESTING OFFICER: POLIZZI,CHRISTINE  
ARRESTING OFFICER AGENCY: Martin County SO  
LAST AGENCY INSPECTION: 10/13/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 31715080A3  
DRY GAS EXPIRY DATE: 12/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 01:21  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: CHOCANO IBARRA

DATE/TIME OF TEST: 10/30/2016 02:07:27

---

Diagnostics Check	OK	02:00
Air Blank	0.000	02:01
Control Test	0.079	02:01
Air Blank	0.000	02:02
Subject Sample # 1	0.174	02:03
Breath Volume	1.140	02:03
Air Blank	0.000	02:03
Air Blank	0.000	02:05
Subject Sample # 2	0.160	02:05
Breath Volume	1.488	02:05
Air Blank	0.000	02:06
Control Test	0.077	02:06
Air Blank	0.000	02:07
Diagnostics Check	OK	02:07

---

Data Download Date/Time: 11/16/2016 14:01

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SINCORE,MICHAEL,A  
BREATH TEST OPERATOR AGENCY: Homestead PD  
ARRESTING OFFICER: PEREZ,RAUL  
ARRESTING OFFICER AGENCY: Homestead PD  
LAST AGENCY INSPECTION: 09/27/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1870563  
DRY GAS EXPIRY DATE: 06/17/2018  
VIOLATION CODE: DUI  
ARREST TIME: 01:12  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: DOMINGO JOSE

DATE/TIME OF TEST: 10/01/2016 02:01:12

---

Diagnostics Check	OK	01:53
Air Blank	0.000	01:54
Control Test	0.081	01:54
Air Blank	0.000	01:55
Subject Sample # 1	0.108	01:56
Breath Volume	1.972	01:56
Air Blank	0.000	01:57
Air Blank	0.000	01:59
Subject Sample # 2	0.101	01:59
Breath Volume	1.714	01:59
Air Blank	0.000	02:00
Control Test	0.080	02:00
Air Blank	0.000	02:01
Diagnostics Check	OK	02:01

---

Data Download Date/Time: 10/27/2016 10:14

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PEREZ,RAUL,  
BREATH TEST OPERATOR AGENCY: Homestead PD  
ARRESTING OFFICER: PEREZ,RAUL  
ARRESTING OFFICER AGENCY: Homestead PD  
LAST AGENCY INSPECTION: 09/27/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1870563  
DRY GAS EXPIRY DATE: 06/17/2018  
VIOLATION CODE: DUI  
ARREST TIME: 02:41  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: PEREZ CARRILLO

DATE/TIME OF TEST: 10/08/2016 03:43:31

---

Diagnostics Check	OK	03:36
Air Blank	0.000	03:37
Control Test	0.081	03:37
Air Blank	0.000	03:38
Subject Sample # 1	0.154	03:38
Breath Volume	3.007	03:38
Air Blank	0.000	03:39
Air Blank	0.000	03:41
Subject Sample # 2	0.153	03:41
Breath Volume	3.105	03:41
Air Blank	0.000	03:42
Control Test	0.080	03:42
Air Blank	0.000	03:43
Diagnostics Check	OK	03:43

---

Data Download Date/Time: 10/27/2016 10:15

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PEREZ,RAUL,  
BREATH TEST OPERATOR AGENCY: Homestead PD  
ARRESTING OFFICER: PEREZ,RAUL  
ARRESTING OFFICER AGENCY: Homestead PD  
LAST AGENCY INSPECTION: 09/27/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1870563  
DRY GAS EXPIRY DATE: 06/17/2018  
VIOLATION CODE: DUI  
ARREST TIME: 01:19  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: GARCIA GODINEZ

DATE/TIME OF TEST: 10/09/2016 02:10:18

---

Diagnostics Check	OK	02:01	
Air Blank	0.000	02:02	
Control Test	0.080	02:02	
Air Blank	0.000	02:03	
Subject Sample	# 1	02:06	No Sample Provided
Breath Volume	0.000	02:06	
Air Blank	0.000	02:06	
Air Blank	0.000	02:08	
Subject Sample	# 2	02:08	Subject Test Refused
Breath Volume	0.000	02:08	
Air Blank	0.000	02:09	
Control Test	0.081	02:09	
Air Blank	0.000	02:10	
Diagnostics Check	OK	02:10	

---

Data Download Date/Time: 10/27/2016 10:15

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PEREZ,RAUL,  
BREATH TEST OPERATOR AGENCY: Homestead PD  
ARRESTING OFFICER: SINCORE,MICHAEL  
ARRESTING OFFICER AGENCY: Homestead PD  
LAST AGENCY INSPECTION: 09/27/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1870563  
DRY GAS EXPIRY DATE: 06/17/2018  
VIOLATION CODE: DUI  
ARREST TIME: 04:36  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:50

SUBJECT'S LAST NAME: TORRES MELO

DATE/TIME OF TEST: 10/10/2016 05:19:44

---

Diagnostics Check	OK	05:11	
Air Blank	0.000	05:11	
Control Test	0.081	05:12	
Air Blank	0.000	05:12	
Subject Sample	# 1	05:15	No Sample Provided
Breath Volume	0.000	05:15	
Air Blank	0.000	05:16	
Air Blank	0.000	05:18	
Subject Sample	# 2	05:18	Subject Test Refused
Breath Volume	0.000	05:18	
Air Blank	0.000	05:18	
Control Test	0.082	05:19	
Air Blank	0.000	05:19	
Diagnostics Check	OK	05:19	

---

Data Download Date/Time: 10/27/2016 10:15

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SINCORE,MICHAEL,A  
BREATH TEST OPERATOR AGENCY: Homestead PD  
ARRESTING OFFICER: SINCORE,MICHAEL  
ARRESTING OFFICER AGENCY: Homestead PD  
LAST AGENCY INSPECTION: 09/27/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1870563  
DRY GAS EXPIRY DATE: 06/17/2018  
VIOLATION CODE: DUI  
ARREST TIME: 05:18  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:24

SUBJECT'S LAST NAME: LEON

DATE/TIME OF TEST: 10/23/2016 06:08:53

---

Diagnostics Check	OK	06:01
Air Blank	0.000	06:02
Control Test	0.081	06:02
Air Blank	0.000	06:03
Subject Sample # 1	0.219	06:03
Breath Volume	2.394	06:03
Air Blank	0.000	06:04
Air Blank	0.000	06:06
Subject Sample # 2	0.231	06:07
Breath Volume	2.757	06:07
Air Blank	0.000	06:07
Control Test	0.079	06:08
Air Blank	0.000	06:08
Diagnostics Check	OK	06:08

---

Data Download Date/Time: 10/27/2016 10:15

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006214

INSTRUMENT REGISTERED TO: OCALA POLICE DEPT  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JOEDICKE,WILLIAM,R  
BREATH TEST OPERATOR AGENCY: Ocala PD  
ARRESTING OFFICER: JOEDICKE,WILLIAM  
ARRESTING OFFICER AGENCY: Ocala PD  
LAST AGENCY INSPECTION: 09/07/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 11315080A1  
DRY GAS EXPIRY DATE: 05/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 01:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:43

SUBJECT'S LAST NAME: CLARK

DATE/TIME OF TEST: 10/06/2016 01:21:20

---

Diagnostics Check	OK	01:14
Air Blank	0.000	01:14
Control Test	0.077	01:15
Air Blank	0.000	01:15
Subject Sample # 1	0.137	01:16
Breath Volume	1.363	01:16
Air Blank	0.000	01:17
Air Blank	0.000	01:18
Subject Sample # 2	0.145	01:19
Breath Volume	2.011	01:19
Air Blank	0.000	01:20
Control Test	0.075	01:20
Air Blank	0.000	01:21
Diagnostics Check	OK	01:21

---

Data Download Date/Time: 10/31/2016 22:40

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006214

INSTRUMENT REGISTERED TO: OCALA POLICE DEPT  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JOEDICKE,WILLIAM,R  
BREATH TEST OPERATOR AGENCY: Ocala PD  
ARRESTING OFFICER: JOEDICKE,WILLIAM  
ARRESTING OFFICER AGENCY: Ocala PD  
LAST AGENCY INSPECTION: 09/07/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 11315080A1  
DRY GAS EXPIRY DATE: 05/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 19:37  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:21

SUBJECT'S LAST NAME: APONTE

DATE/TIME OF TEST: 10/06/2016 19:47:18

---

Diagnostics Check	OK	19:41
Air Blank	0.000	19:41
Control Test	0.077	19:41
Air Blank	0.000	19:42
Subject Sample # 1	0.124	19:42
Breath Volume	3.214	19:42
Air Blank	0.000	19:43
Air Blank	0.000	19:45
Subject Sample # 2	0.123	19:45
Breath Volume	3.472	19:45
Air Blank	0.000	19:46
Control Test	0.076	19:46
Air Blank	0.000	19:47
Diagnostics Check	OK	19:47

---

Data Download Date/Time: 10/31/2016 22:40



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006214

INSTRUMENT REGISTERED TO: OCALA POLICE DEPT  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JOEDICKE,WILLIAM,R  
BREATH TEST OPERATOR AGENCY: Ocala PD  
ARRESTING OFFICER: JOEDICKE,WILLIAM  
ARRESTING OFFICER AGENCY: Ocala PD  
LAST AGENCY INSPECTION: 09/07/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 11315080A1  
DRY GAS EXPIRY DATE: 05/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 23:20  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:03

SUBJECT'S LAST NAME: CARR

DATE/TIME OF TEST: 10/06/2016 23:33:13

---

Diagnostics Check	OK	23:26	
Air Blank	0.000	23:27	
Control Test	0.076	23:27	
Air Blank	0.000	23:27	
Subject Sample	# 1 0.182	23:28	
Breath Volume	1.843	23:28	
Air Blank	0.000	23:29	
Air Blank	0.000	23:31	
Subject Sample	# 2 0.185	23:31	
Breath Volume	2.703	23:31	
Air Blank	0.000	23:32	
Control Test	0.074	23:32	Control Outside Tolerance
Air Blank	0.000	23:33	

---

Data Download Date/Time: 10/31/2016 22:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006214

INSTRUMENT REGISTERED TO: OCALA POLICE DEPT  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JOEDICKE,WILLIAM,R  
BREATH TEST OPERATOR AGENCY: Ocala PD  
ARRESTING OFFICER: JOEDICKE,WILLIAM  
ARRESTING OFFICER AGENCY: Ocala PD  
LAST AGENCY INSPECTION: 09/07/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 11315080A1  
DRY GAS EXPIRY DATE: 05/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 23:20  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:03

SUBJECT'S LAST NAME: CARR

DATE/TIME OF TEST: 10/06/2016 23:41:30

---

Diagnostics Check	OK	23:35
Air Blank	0.000	23:35
Control Test	0.075	23:36
Air Blank	0.000	23:36
Subject Sample # 1	0.179	23:37
Breath Volume	2.566	23:37
Air Blank	0.000	23:37
Air Blank	0.000	23:39
Subject Sample # 2	0.181	23:40
Breath Volume	2.839	23:40
Air Blank	0.000	23:40
Control Test	0.075	23:40
Air Blank	0.000	23:41
Diagnostics Check	OK	23:41

---

Data Download Date/Time: 10/31/2016 22:40

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006214

INSTRUMENT REGISTERED TO: OCALA POLICE DEPT  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JOEDICKE,WILLIAM,R  
BREATH TEST OPERATOR AGENCY: Ocala PD  
ARRESTING OFFICER: JOEDICKE,WILLIAM  
ARRESTING OFFICER AGENCY: Ocala PD  
LAST AGENCY INSPECTION: 09/07/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 11315080A1  
DRY GAS EXPIRY DATE: 05/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 20:13  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:58

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 10/08/2016 20:26:07

---

Diagnostics Check	OK	20:18	
Air Blank	0.000	20:19	
Control Test	0.077	20:19	
Air Blank	0.000	20:20	
Subject Sample	# 1 0.267	20:20	
Breath Volume	1.246	20:20	
Air Blank	0.000	20:21	
Air Blank	0.000	20:23	
Subject Sample	# 2 0.269	20:24	
Breath Volume	1.468	20:24	
Air Blank	0.000	20:25	
Control Test	0.074	20:25	Control Outside Tolerance
Air Blank	0.000	20:26	

---

Data Download Date/Time: 10/31/2016 22:40

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006214

INSTRUMENT REGISTERED TO: OCALA POLICE DEPT  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JOEDICKE,WILLIAM,R  
BREATH TEST OPERATOR AGENCY: Ocala PD  
ARRESTING OFFICER: JOEDICKE,WILLIAM  
ARRESTING OFFICER AGENCY: Ocala PD  
LAST AGENCY INSPECTION: 09/07/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 11315080A1  
DRY GAS EXPIRY DATE: 05/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 20:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:57

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 10/08/2016 20:33:04

---

Diagnostics Check	OK	20:27	
Air Blank	0.000	20:28	
Control Test	0.076	20:28	
Air Blank	0.000	20:29	
Subject Sample # 1	0.264	20:29	
Breath Volume	1.132	20:29	
Air Blank	0.000	20:30	
Air Blank		20:31	RFI Detect
Air Blank	0.000	20:32	
Control Test	0.076	20:32	
Air Blank	0.000	20:32	
Diagnostics Check	OK	20:33	

---

Data Download Date/Time: 10/31/2016 22:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006214

INSTRUMENT REGISTERED TO: OCALA POLICE DEPT  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JOEDICKE,WILLIAM,R  
BREATH TEST OPERATOR AGENCY: Ocala PD  
ARRESTING OFFICER: JOEDICKE,WILLIAM  
ARRESTING OFFICER AGENCY: Ocala PD  
LAST AGENCY INSPECTION: 09/07/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 11315080A1  
DRY GAS EXPIRY DATE: 05/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 20:13  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:57

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 10/08/2016 20:39:58

---

Diagnostics Check	OK	20:34	
Air Blank	0.000	20:34	
Control Test	0.077	20:35	
Air Blank	0.000	20:35	
Subject Sample	# 1 0.250	20:37	
Breath Volume	1.164	20:37	
Air Blank	0.000	20:37	
Air Blank		20:39	Ambient Fail
Air Blank	0.000	20:39	

---

Data Download Date/Time: 10/31/2016 22:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006214

INSTRUMENT REGISTERED TO: OCALA POLICE DEPT  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JOEDICKE,WILLIAM,R  
BREATH TEST OPERATOR AGENCY: Ocala PD  
ARRESTING OFFICER: JOEDICKE,WILLIAM  
ARRESTING OFFICER AGENCY: Ocala PD  
LAST AGENCY INSPECTION: 09/07/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 11315080A1  
DRY GAS EXPIRY DATE: 05/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 20:13  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:57

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 10/08/2016 20:47:55

---

Diagnostics Check	OK	20:41
Air Blank	0.000	20:42
Control Test	0.076	20:42
Air Blank	0.000	20:42
Subject Sample # 1	0.253	20:43
Breath Volume	1.613	20:43
Air Blank	0.000	20:44
Air Blank	0.000	20:45
Subject Sample # 2	0.262	20:46
Breath Volume	1.667	20:46
Air Blank	0.000	20:46
Control Test	0.076	20:47
Air Blank	0.000	20:47
Diagnostics Check	OK	20:47

---

Data Download Date/Time: 10/31/2016 22:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006214

INSTRUMENT REGISTERED TO: OCALA POLICE DEPT  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JOEDICKE,WILLIAM,R  
BREATH TEST OPERATOR AGENCY: Ocala PD  
ARRESTING OFFICER: JOEDICKE,WILLIAM  
ARRESTING OFFICER AGENCY: Ocala PD  
LAST AGENCY INSPECTION: 09/07/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 11315080A1  
DRY GAS EXPIRY DATE: 05/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 00:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: HYATT

DATE/TIME OF TEST: 10/20/2016 00:59:16

---

Diagnostics Check	OK	00:52
Air Blank	0.000	00:53
Control Test	0.078	00:53
Air Blank	0.000	00:53
Subject Sample # 1	0.196	00:54
Breath Volume	3.613	00:54
Air Blank	0.000	00:55
Air Blank	0.000	00:57
Subject Sample # 2	0.186	00:57
Breath Volume	2.812	00:57
Air Blank	0.000	00:58
Control Test	0.076	00:58
Air Blank	0.000	00:59
Diagnostics Check	OK	00:59

---

Data Download Date/Time: 10/31/2016 22:40

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006214

INSTRUMENT REGISTERED TO: OCALA POLICE DEPT  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JOEDICKE,WILLIAM,R  
BREATH TEST OPERATOR AGENCY: Ocala PD  
ARRESTING OFFICER: JOEDICKE,WILLIAM  
ARRESTING OFFICER AGENCY: Ocala PD  
LAST AGENCY INSPECTION: 09/07/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 11315080A1  
DRY GAS EXPIRY DATE: 05/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 01:09  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:43

SUBJECT'S LAST NAME: SAGASTUME SARMIENTO

DATE/TIME OF TEST: 10/22/2016 01:29:58

---

Diagnostics Check		OK	01:14	
Air Blank		0.000	01:14	
Control Test		0.081	01:15	
Air Blank		0.000	01:15	
Subject Sample	# 1	0.171	01:18	Volume Not Met
Breath Volume		0.226	01:18	
Air Blank		0.000	01:19	
Air Blank		0.000	01:21	
Subject Sample	# 2	0.168	01:22	
Breath Volume		1.218	01:22	
Air Blank		0.000	01:23	
Air Blank		0.000	01:25	
Subject Sample	# 3	0.158	01:28	Volume Not Met
Breath Volume		0.585	01:28	
Air Blank		0.000	01:29	
Control Test		0.078	01:29	
Air Blank		0.000	01:29	
Diagnostics Check		OK	01:29	

---

Data Download Date/Time: 10/31/2016 22:40



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006214

INSTRUMENT REGISTERED TO: OCALA POLICE DEPT  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JOEDICKE,WILLIAM,R  
BREATH TEST OPERATOR AGENCY: Ocala PD  
ARRESTING OFFICER: JOEDICKE,WILLIAM  
ARRESTING OFFICER AGENCY: Ocala PD  
LAST AGENCY INSPECTION: 09/07/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 11315080A1  
DRY GAS EXPIRY DATE: 05/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 01:09  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:43

SUBJECT'S LAST NAME: SAGASTUME SARMIENTO

DATE/TIME OF TEST: 10/22/2016 01:39:08

---

Diagnostics Check	OK	01:31
Air Blank	0.000	01:31
Control Test	0.079	01:32
Air Blank	0.000	01:32
Subject Sample # 1	0.169	01:33
Breath Volume	2.097	01:33
Air Blank	0.000	01:33
Air Blank	0.000	01:35
Subject Sample # 2	0.169	01:37
Breath Volume	1.300	01:37
Air Blank	0.000	01:38
Control Test	0.079	01:38
Air Blank	0.000	01:39
Diagnostics Check	OK	01:39

---

Data Download Date/Time: 10/31/2016 22:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006221

INSTRUMENT REGISTERED TO: LAKE CITY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HIGHTOWER,CONNIE,A  
BREATH TEST OPERATOR AGENCY: Lake City PD  
ARRESTING OFFICER: HIGHTOWER,CONNIE  
ARRESTING OFFICER AGENCY: Lake City PD  
LAST AGENCY INSPECTION: 10/09/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 477627  
DRY GAS EXPIRY DATE: 06/29/2019  
VIOLATION CODE: DUI  
ARREST TIME: 18:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:57

SUBJECT'S LAST NAME: COOPER

DATE/TIME OF TEST: 10/23/2016 19:29:21

---

Diagnostics Check	OK	19:22
Air Blank	0.000	19:23
Control Test	0.080	19:23
Air Blank	0.000	19:24
Subject Sample # 1	0.259	19:24
Breath Volume	1.796	19:24
Air Blank	0.000	19:25
Air Blank	0.000	19:27
Subject Sample # 2	0.267	19:27
Breath Volume	2.265	19:27
Air Blank	0.000	19:28
Control Test	0.080	19:28
Air Blank	0.000	19:29
Diagnostics Check	OK	19:29

---

Data Download Date/Time: 11/06/2016 23:32

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CASE, THOMAS, D  
BREATH TEST OPERATOR AGENCY: Brevard County SO  
ARRESTING OFFICER: BARBER, AMANDA  
ARRESTING OFFICER AGENCY: Brevard County SO  
LAST AGENCY INSPECTION: 09/21/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 12716080A5  
DRY GAS EXPIRY DATE: 06/05/2018  
VIOLATION CODE: PROB/PAROL  
ARREST TIME: 22:24  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:24

SUBJECT'S LAST NAME: STAPLETON

DATE/TIME OF TEST: 10/02/2016 23:05:02

---

Diagnostics Check	OK	23:01	
Air Blank	0.000	23:01	
Control Test	0.079	23:01	
Air Blank	0.000	23:02	
Subject Sample	# 1	23:04	Sequence Aborted
Breath Volume	0.000	23:04	
Air Blank	0.000	23:05	

---

Data Download Date/Time: 10/19/2016 20:18

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILKIN,SEAN,D  
BREATH TEST OPERATOR AGENCY: Brevard County SO  
ARRESTING OFFICER: BARRETT,TAYLOR  
ARRESTING OFFICER AGENCY: Brevard County SO  
LAST AGENCY INSPECTION: 09/21/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 12716080A5  
DRY GAS EXPIRY DATE: 06/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 02:01  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:01

SUBJECT'S LAST NAME: ZITTEL

DATE/TIME OF TEST: 10/04/2016 02:29:20

---

Diagnostics Check		OK	02:21
Air Blank		0.000	02:22
Control Test		0.079	02:22
Air Blank		0.000	02:22
Subject Sample	# 1	0.206	02:24
Breath Volume		2.472	02:24
Air Blank		0.000	02:25
Air Blank		0.000	02:27
Subject Sample	# 2	0.194	02:27
Breath Volume		1.578	02:27
Air Blank		0.000	02:28
Control Test		0.077	02:28
Air Blank		0.000	02:29
Diagnostics Check		OK	02:29

---

Data Download Date/Time: 10/19/2016 20:18

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BIVINS, HEATHER, C  
BREATH TEST OPERATOR AGENCY: Brevard County SO  
ARRESTING OFFICER: MOORE, KIMBERLY  
ARRESTING OFFICER AGENCY: Brevard County SO  
LAST AGENCY INSPECTION: 09/21/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 12716080A5  
DRY GAS EXPIRY DATE: 06/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 19:42  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: HARRINGTON

DATE/TIME OF TEST: 10/12/2016 20:28:56

---

Diagnostics Check	OK	20:21
Air Blank	0.000	20:22
Control Test	0.077	20:22
Air Blank	0.000	20:22
Subject Sample # 1	0.131	20:24
Breath Volume	1.515	20:24
Air Blank	0.000	20:24
Air Blank	0.000	20:26
Subject Sample # 2	0.139	20:27
Breath Volume	1.421	20:27
Air Blank	0.000	20:28
Control Test	0.077	20:28
Air Blank	0.000	20:28
Diagnostics Check	OK	20:28

---

Data Download Date/Time: 10/19/2016 20:18

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SCHEINER,JAMES,R  
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD  
ARRESTING OFFICER: SCHEINER,JAMES  
ARRESTING OFFICER AGENCY: Cocoa Beach PD  
LAST AGENCY INSPECTION: 09/21/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 12716080A5  
DRY GAS EXPIRY DATE: 06/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 02:38  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: ESTRADA LARROSA

DATE/TIME OF TEST: 10/14/2016 03:19:26

---

Diagnostics Check	OK	03:12
Air Blank	0.000	03:12
Control Test	0.077	03:13
Air Blank	0.000	03:13
Subject Sample # 1	0.146	03:14
Breath Volume	1.386	03:14
Air Blank	0.000	03:14
Air Blank	0.000	03:16
Subject Sample # 2	0.159	03:17
Breath Volume	1.812	03:17
Air Blank	0.000	03:18
Control Test	0.077	03:18
Air Blank	0.000	03:19
Diagnostics Check	OK	03:19

---

Data Download Date/Time: 10/19/2016 20:18

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BIVINS,HEATHER,C  
BREATH TEST OPERATOR AGENCY: Brevard County SO  
ARRESTING OFFICER: BIVINS,HEATHER  
ARRESTING OFFICER AGENCY: Brevard County SO  
LAST AGENCY INSPECTION: 10/19/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 12716080A5  
DRY GAS EXPIRY DATE: 06/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 02:41  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:51

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 10/21/2016 03:22:03

---

Diagnostics Check	OK	03:13
Air Blank	0.000	03:13
Control Test	0.078	03:14
Air Blank	0.000	03:14
Subject Sample # 1	0.158	03:17
Breath Volume	3.238	03:17
Air Blank	0.000	03:17
Air Blank	0.000	03:19
Subject Sample # 2	0.163	03:20
Breath Volume	1.191	03:20
Air Blank	0.000	03:21
Control Test	0.077	03:21
Air Blank	0.000	03:21
Diagnostics Check	OK	03:22

---

Data Download Date/Time: 11/16/2016 18:34

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BIVINS,HEATHER,C  
BREATH TEST OPERATOR AGENCY: Brevard County SO  
ARRESTING OFFICER: BIVINS,HEATHER  
ARRESTING OFFICER AGENCY: Brevard County SO  
LAST AGENCY INSPECTION: 10/19/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 12716080A5  
DRY GAS EXPIRY DATE: 06/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 01:48  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: SMITS

DATE/TIME OF TEST: 10/30/2016 02:24:58

---

Diagnostics Check	OK	02:17
Air Blank	0.000	02:18
Control Test	0.078	02:18
Air Blank	0.000	02:19
Subject Sample # 1	0.249	02:20
Breath Volume	3.707	02:20
Air Blank	0.000	02:20
Air Blank	0.000	02:22
Subject Sample # 2	0.248	02:23
Breath Volume	2.718	02:23
Air Blank	0.000	02:24
Control Test	0.078	02:24
Air Blank	0.000	02:24
Diagnostics Check	OK	02:24

---

Data Download Date/Time: 11/16/2016 18:34



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006234

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CAMERON,DAVID,J  
BREATH TEST OPERATOR AGENCY: Hendry County SO  
ARRESTING OFFICER: CAMERON,DAVID  
ARRESTING OFFICER AGENCY: Hendry County SO  
LAST AGENCY INSPECTION: 10/26/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1694353  
DRY GAS EXPIRY DATE: 06/10/2017  
VIOLATION CODE: DUI  
ARREST TIME: 23:48  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:04

SUBJECT'S LAST NAME: PELICO

DATE/TIME OF TEST: 10/29/2016 00:30:52

---

Diagnostics Check	OK	00:24
Air Blank	0.000	00:24
Control Test	0.081	00:24
Air Blank	0.000	00:25
Subject Sample # 1	0.178	00:26
Breath Volume	2.652	00:26
Air Blank	0.000	00:27
Air Blank	0.000	00:28
Subject Sample # 2	0.190	00:29
Breath Volume	2.785	00:29
Air Blank	0.000	00:29
Control Test	0.080	00:30
Air Blank	0.000	00:30
Diagnostics Check	OK	00:30

---

Data Download Date/Time: 11/28/2016 20:26

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROWE,BRANDON,D  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: ROWE,BRANDON  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 09/26/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 203498  
DRY GAS EXPIRY DATE: 12/10/2018  
VIOLATION CODE: DUI  
ARREST TIME: 01:44  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:47

SUBJECT'S LAST NAME: DAY

DATE/TIME OF TEST: 10/04/2016 02:17:34

---

Diagnostics Check	OK	02:11
Air Blank	0.000	02:11
Control Test	0.081	02:11
Air Blank	0.000	02:12
Subject Sample # 1	0.208	02:12
Breath Volume	3.054	02:12
Air Blank	0.000	02:13
Air Blank	0.000	02:15
Subject Sample # 2	0.210	02:15
Breath Volume	2.539	02:15
Air Blank	0.000	02:16
Control Test	0.080	02:16
Air Blank	0.000	02:17
Diagnostics Check	OK	02:17

---

Data Download Date/Time: 10/17/2016 12:21

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROWE,BRANDON,D  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: ROWE,BRANDON  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 09/26/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 203498  
DRY GAS EXPIRY DATE: 12/10/2018  
VIOLATION CODE: DUI  
ARREST TIME: 20:46  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 10/08/2016 21:19:39

---

Diagnostics Check	OK	21:11
Air Blank	0.000	21:11
Control Test	0.080	21:12
Air Blank	0.000	21:12
Subject Sample # 1	0.000	21:15
Breath Volume	2.269	21:15
Air Blank	0.000	21:15
Air Blank	0.000	21:17
Subject Sample # 2	0.000	21:18
Breath Volume	1.554	21:18
Air Blank	0.000	21:18
Control Test	0.079	21:19
Air Blank	0.000	21:19
Diagnostics Check	OK	21:19

---

Data Download Date/Time: 10/17/2016 12:21

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROWE,BRANDON,D  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: ROWE,BRANDON  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 10/17/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 203498  
DRY GAS EXPIRY DATE: 12/10/2018  
VIOLATION CODE: DUI  
ARREST TIME: 02:04  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:09

SUBJECT'S LAST NAME: SOTO ESCALANTE

DATE/TIME OF TEST: 10/29/2016 02:39:20

---

Diagnostics Check	OK	02:30
Air Blank	0.000	02:30
Control Test	0.080	02:31
Air Blank	0.000	02:31
Subject Sample # 1	0.119	02:34
Breath Volume	1.507	02:34
Air Blank	0.000	02:35
Air Blank	0.000	02:37
Subject Sample # 2	0.116	02:37
Breath Volume	1.156	02:37
Air Blank	0.000	02:38
Control Test	0.080	02:38
Air Blank	0.000	02:39
Diagnostics Check	OK	02:39

---

Data Download Date/Time: 11/28/2016 09:22

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: STACY,A,L  
BREATH TEST OPERATOR AGENCY: Clay County SO  
ARRESTING OFFICER: SAUNDERS,J  
ARRESTING OFFICER AGENCY: Clay County SO  
LAST AGENCY INSPECTION: 09/25/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24515080A3  
DRY GAS EXPIRY DATE: 10/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 16:56  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:12

SUBJECT'S LAST NAME: WATSON

DATE/TIME OF TEST: 10/06/2016 17:38:57

---

Diagnostics Check	OK	17:32
Air Blank	0.000	17:33
Control Test	0.080	17:33
Air Blank	0.000	17:34
Subject Sample # 1	0.176	17:34
Breath Volume	2.316	17:34
Air Blank	0.000	17:35
Air Blank	0.000	17:36
Subject Sample # 2	0.173	17:37
Breath Volume	2.132	17:37
Air Blank	0.000	17:38
Control Test	0.079	17:38
Air Blank	0.000	17:38
Diagnostics Check	OK	17:38

---

Data Download Date/Time: 10/09/2016 09:41

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ZIER,TAMMY,L  
BREATH TEST OPERATOR AGENCY: Clay County SO  
ARRESTING OFFICER: GADT,MARC  
ARRESTING OFFICER AGENCY: Clay County SO  
LAST AGENCY INSPECTION: 10/09/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24515080A3  
DRY GAS EXPIRY DATE: 10/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 18:32  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 19:50

SUBJECT'S LAST NAME: CRISWELL

DATE/TIME OF TEST: 10/12/2016 20:17:16

---

Diagnostics Check	OK	20:10
Air Blank	0.000	20:11
Control Test	0.081	20:11
Air Blank	0.000	20:12
Subject Sample # 1	0.032	20:12
Breath Volume	2.078	20:12
Air Blank	0.000	20:13
Air Blank	0.000	20:15
Subject Sample # 2	0.031	20:15
Breath Volume	2.152	20:15
Air Blank	0.000	20:16
Control Test	0.080	20:16
Air Blank	0.000	20:17
Diagnostics Check	OK	20:17

---

Data Download Date/Time: 11/06/2016 12:04

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: COE,J,D  
BREATH TEST OPERATOR AGENCY: Clay County SO  
ARRESTING OFFICER: SAUNDERS,J  
ARRESTING OFFICER AGENCY: Clay County SO  
LAST AGENCY INSPECTION: 10/09/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24515080A3  
DRY GAS EXPIRY DATE: 10/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 16:16  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:42

SUBJECT'S LAST NAME: HERSEY

DATE/TIME OF TEST: 10/14/2016 17:49:40

---

Diagnostics Check	OK	17:46	
Air Blank	0.000	17:47	
Control Test	0.080	17:47	
Air Blank	0.000	17:48	
Subject Sample	# 1	17:48	Subject Test Refused
Breath Volume	0.000	17:48	
Air Blank	0.000	17:48	
Control Test	0.080	17:49	
Air Blank	0.000	17:49	
Diagnostics Check	OK	17:49	

---

Data Download Date/Time: 11/06/2016 12:04

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ZIER,TAMMY,L  
BREATH TEST OPERATOR AGENCY: Clay County SO  
ARRESTING OFFICER: WRIGHT,BRYAN  
ARRESTING OFFICER AGENCY: Clay County SO  
LAST AGENCY INSPECTION: 10/09/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24515080A3  
DRY GAS EXPIRY DATE: 10/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 02:06  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:37

SUBJECT'S LAST NAME: HOLMES

DATE/TIME OF TEST: 10/22/2016 03:07:39

---

Diagnostics Check	OK	02:58	
Air Blank	0.000	02:59	
Control Test	0.081	02:59	
Air Blank	0.000	02:59	
Subject Sample	# 1	03:03	No Sample Provided
Breath Volume	0.000	03:03	
Air Blank	0.000	03:03	
Air Blank	0.000	03:05	
Subject Sample	# 2	03:06	Subject Test Refused
Breath Volume	0.000	03:06	
Air Blank	0.000	03:06	
Control Test	0.081	03:07	
Air Blank	0.000	03:07	
Diagnostics Check	OK	03:07	

---

Data Download Date/Time: 11/06/2016 12:04



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JOHONSON, MATTHEW, P  
BREATH TEST OPERATOR AGENCY: Clay County SO  
ARRESTING OFFICER: WHITE, M  
ARRESTING OFFICER AGENCY: Clay County SO  
LAST AGENCY INSPECTION: 10/09/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24515080A3  
DRY GAS EXPIRY DATE: 10/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 03:23  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:01

SUBJECT'S LAST NAME: MURRAY

DATE/TIME OF TEST: 10/22/2016 04:28:35

---

Diagnostics Check	OK	04:22
Air Blank	0.000	04:22
Control Test	0.080	04:22
Air Blank	0.000	04:23
Subject Sample # 1	0.146	04:23
Breath Volume	2.113	04:23
Air Blank	0.000	04:24
Air Blank	0.000	04:26
Subject Sample # 2	0.158	04:26
Breath Volume	2.761	04:26
Air Blank	0.000	04:27
Control Test	0.079	04:27
Air Blank	0.000	04:28
Diagnostics Check	OK	04:28

---

Data Download Date/Time: 11/06/2016 12:04

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ASHE,KEVIN,B  
BREATH TEST OPERATOR AGENCY: Clay County SO  
ARRESTING OFFICER: ALEXANDER,D  
ARRESTING OFFICER AGENCY: Clay County SO  
LAST AGENCY INSPECTION: 10/09/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24515080A3  
DRY GAS EXPIRY DATE: 10/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 21:50  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:26

SUBJECT'S LAST NAME: BYRD

DATE/TIME OF TEST: 10/22/2016 22:54:59

---

Diagnostics Check	OK	22:47
Air Blank	0.000	22:47
Control Test	0.081	22:47
Air Blank	0.000	22:48
Subject Sample # 1	0.122	22:49
Breath Volume	2.000	22:49
Air Blank	0.000	22:50
Air Blank	0.000	22:52
Subject Sample # 2	0.123	22:53
Breath Volume	2.351	22:53
Air Blank	0.000	22:54
Control Test	0.080	22:54
Air Blank	0.000	22:54
Diagnostics Check	OK	22:54

---

Data Download Date/Time: 11/06/2016 12:04

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ZIER,TAMMY,L  
BREATH TEST OPERATOR AGENCY: Clay County SO  
ARRESTING OFFICER: GLENN,CHRIS  
ARRESTING OFFICER AGENCY: Clay County SO  
LAST AGENCY INSPECTION: 10/09/2016

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 24515080A3  
DRY GAS EXPIRY DATE: 10/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 22:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: LEISS

DATE/TIME OF TEST: 10/23/2016 00:48:05

---

Diagnostics Check		OK	00:41
Air Blank		0.000	00:41
Control Test		0.080	00:41
Air Blank		0.000	00:42
Subject Sample	# 1	0.209	00:43
Breath Volume		1.687	00:43
Air Blank		0.000	00:43
Air Blank		0.000	00:45
Subject Sample	# 2	0.209	00:46
Breath Volume		1.964	00:46
Air Blank		0.000	00:47
Control Test		0.082	00:47
Air Blank		0.000	00:48
Diagnostics Check		OK	00:48

---

Data Download Date/Time: 11/06/2016 12:04

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JONES,DEPUTY,T  
BREATH TEST OPERATOR AGENCY: Clay County SO  
ARRESTING OFFICER: GRACKI,DEPUTY  
ARRESTING OFFICER AGENCY: Clay County SO  
LAST AGENCY INSPECTION: 10/09/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24515080A3  
DRY GAS EXPIRY DATE: 10/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 04:37  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:50

SUBJECT'S LAST NAME: SANTOS

DATE/TIME OF TEST: 10/23/2016 06:19:47

---

Diagnostics Check	OK	06:13
Air Blank	0.000	06:13
Control Test	0.081	06:13
Air Blank	0.000	06:14
Subject Sample # 1	0.069	06:15
Breath Volume	1.847	06:15
Air Blank	0.000	06:15
Air Blank	0.000	06:17
Subject Sample # 2	0.072	06:18
Breath Volume	1.937	06:18
Air Blank	0.000	06:18
Control Test	0.080	06:19
Air Blank	0.000	06:19
Diagnostics Check	OK	06:19

---

Data Download Date/Time: 11/06/2016 12:04

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RANDOPH,KEVIN,J  
BREATH TEST OPERATOR AGENCY: Clay County SO  
ARRESTING OFFICER: SAUNDERS,J  
ARRESTING OFFICER AGENCY: Clay County SO  
LAST AGENCY INSPECTION: 10/09/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24515080A3  
DRY GAS EXPIRY DATE: 10/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 14:33  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 15:13

SUBJECT'S LAST NAME: VANBOSKKERCK

DATE/TIME OF TEST: 10/30/2016 15:37:29

---

Diagnostics Check	OK	15:34	
Air Blank	0.000	15:34	
Control Test	0.080	15:35	
Air Blank	0.000	15:35	
Subject Sample	# 1	15:36	Subject Test Refused
Breath Volume	0.000	15:36	
Air Blank	0.000	15:36	
Control Test	0.081	15:36	
Air Blank	0.000	15:37	
Diagnostics Check	OK	15:37	

---

Data Download Date/Time: 11/06/2016 12:04

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RANDOLPH,KEVIN,J  
BREATH TEST OPERATOR AGENCY: Clay County SO  
ARRESTING OFFICER: SAUNDERS,J  
ARRESTING OFFICER AGENCY: Clay County SO  
LAST AGENCY INSPECTION: 10/09/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24515080A3  
DRY GAS EXPIRY DATE: 10/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 14:33  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 15:13

SUBJECT'S LAST NAME: VANBOSKERCK

DATE/TIME OF TEST: 10/30/2016 15:48:26

---

Diagnostics Check	OK	15:39	
Air Blank	0.000	15:40	
Control Test	0.080	15:40	
Air Blank	0.000	15:41	
Subject Sample	# 1	15:44	No Sample Provided
Breath Volume	0.000	15:44	
Air Blank	0.000	15:44	
Air Blank	0.000	15:46	
Subject Sample	# 2	15:47	Subject Test Refused
Breath Volume	0.000	15:47	
Air Blank	0.000	15:47	
Control Test	0.079	15:47	
Air Blank	0.000	15:48	
Diagnostics Check	OK	15:48	

---

Data Download Date/Time: 11/06/2016 12:04

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: O'NEAL,SHARIL  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: SCARTOZZI,PATRICK  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 09/09/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 279870  
DRY GAS EXPIRY DATE: 02/04/2019  
VIOLATION CODE: DUI  
ARREST TIME: 03:01  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:17

SUBJECT'S LAST NAME: TOOLSEE

DATE/TIME OF TEST: 10/01/2016 03:48:55

---

Diagnostics Check	OK	03:42
Air Blank	0.000	03:43
Control Test	0.081	03:43
Air Blank	0.000	03:43
Subject Sample # 1	0.121	03:44
Breath Volume	2.273	03:44
Air Blank	0.000	03:45
Air Blank	0.000	03:46
Subject Sample # 2	0.117	03:47
Breath Volume	1.308	03:47
Air Blank	0.000	03:47
Control Test	0.081	03:48
Air Blank	0.000	03:48
Diagnostics Check	OK	03:48

---

Data Download Date/Time: 10/14/2016 16:04

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: O'NEAL,SHARIL  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: WARD,CHRISTOPHER  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 09/09/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 279870  
DRY GAS EXPIRY DATE: 02/04/2019  
VIOLATION CODE: DUI  
ARREST TIME: 02:11  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: BANKS

DATE/TIME OF TEST: 10/02/2016 03:37:44

---

Diagnostics Check	OK	03:30
Air Blank	0.000	03:31
Control Test	0.081	03:31
Air Blank	0.000	03:32
Subject Sample # 1	0.218	03:32
Breath Volume	1.527	03:32
Air Blank	0.000	03:33
Air Blank	0.000	03:35
Subject Sample # 2	0.220	03:36
Breath Volume	1.570	03:36
Air Blank	0.000	03:36
Control Test	0.081	03:37
Air Blank	0.000	03:37
Diagnostics Check	OK	03:37

---

Data Download Date/Time: 10/14/2016 16:04



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: OWEN,SUE,  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: WALTON,THOMAS  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 09/09/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 279870  
DRY GAS EXPIRY DATE: 02/04/2019  
VIOLATION CODE: DUI  
ARREST TIME: 21:49  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:48

SUBJECT'S LAST NAME: PALACIOS

DATE/TIME OF TEST: 10/02/2016 23:18:53

---

Diagnostics Check	OK	23:12
Air Blank	0.000	23:13
Control Test	0.081	23:13
Air Blank	0.000	23:14
Subject Sample # 1	0.198	23:14
Breath Volume	1.328	23:14
Air Blank	0.000	23:15
Air Blank	0.000	23:16
Subject Sample # 2	0.196	23:17
Breath Volume	1.925	23:17
Air Blank	0.000	23:17
Control Test	0.081	23:18
Air Blank	0.000	23:18
Diagnostics Check	OK	23:18

---

Data Download Date/Time: 10/14/2016 16:04

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SCHAFFER,NANCY,M  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: BROWN,M  
ARRESTING OFFICER AGENCY: Riviera Beach PD  
LAST AGENCY INSPECTION: 09/09/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 279870  
DRY GAS EXPIRY DATE: 02/04/2019  
VIOLATION CODE: DUI  
ARREST TIME: 13:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 14:20

SUBJECT'S LAST NAME: BRACKNEY

DATE/TIME OF TEST: 10/04/2016 15:02:46

---

Diagnostics Check	OK	14:54
Air Blank	0.000	14:55
Control Test	0.082	14:55
Air Blank	0.000	14:55
Subject Sample # 1	0.283	14:57
Breath Volume	1.539	14:57
Air Blank	0.000	14:58
Air Blank	0.000	15:00
Subject Sample # 2	0.274	15:01
Breath Volume	1.531	15:01
Air Blank	0.000	15:01
Control Test	0.079	15:02
Air Blank	0.000	15:02
Diagnostics Check	OK	15:02

---

Data Download Date/Time: 10/14/2016 16:04

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: O'NEAL,SHARIL  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: FREY,JACOB  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 09/09/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 279870  
DRY GAS EXPIRY DATE: 02/04/2019  
VIOLATION CODE: DUI  
ARREST TIME: 22:33  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: PHILIPPE

DATE/TIME OF TEST: 10/05/2016 23:59:41

---

Diagnostics Check	OK	23:53
Air Blank	0.000	23:54
Control Test	0.082	23:54
Air Blank	0.000	23:54
Subject Sample # 1	0.145	23:55
Breath Volume	1.683	23:55
Air Blank	0.000	23:55
Air Blank	0.000	23:57
Subject Sample # 2	0.147	23:58
Breath Volume	1.914	23:58
Air Blank	0.000	23:58
Control Test	0.081	23:59
Air Blank	0.000	23:59
Diagnostics Check	OK	23:59

---

Data Download Date/Time: 10/14/2016 16:04

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PARENT,GARY,J  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: SENTMANAT,ADOLFO  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 09/09/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 279870  
DRY GAS EXPIRY DATE: 02/04/2019  
VIOLATION CODE: DUI  
ARREST TIME: 02:11  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: REYES-ALVAREZ

DATE/TIME OF TEST: 10/08/2016 03:03:39

---

Diagnostics Check	OK	02:57
Air Blank	0.000	02:57
Control Test	0.082	02:57
Air Blank	0.000	02:58
Subject Sample # 1	0.200	02:59
Breath Volume	2.992	02:59
Air Blank	0.000	02:59
Air Blank	0.000	03:01
Subject Sample # 2	0.198	03:02
Breath Volume	2.667	03:02
Air Blank	0.000	03:02
Control Test	0.081	03:03
Air Blank	0.000	03:03
Diagnostics Check	OK	03:03

---

Data Download Date/Time: 10/14/2016 16:04

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PARENT,GARY,J  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: SENTMANAT,ADOLFO  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 09/09/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 279870  
DRY GAS EXPIRY DATE: 02/04/2019  
VIOLATION CODE: DUI  
ARREST TIME: 00:37  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:56

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 10/09/2016 01:32:30

---

Diagnostics Check	OK	01:23
Air Blank	0.000	01:24
Control Test	0.081	01:24
Air Blank	0.000	01:25
Subject Sample # 1	0.184	01:27
Breath Volume	1.261	01:27
Air Blank	0.000	01:28
Air Blank	0.000	01:29
Subject Sample # 2	0.187	01:30
Breath Volume	1.996	01:30
Air Blank	0.000	01:31
Control Test	0.081	01:31
Air Blank	0.000	01:32
Diagnostics Check	OK	01:32

---

Data Download Date/Time: 10/14/2016 16:04

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CAIN,JEANETTE,  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: CHIQUITO-RODRIGUEZ,LENNDY  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 09/09/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 279870  
DRY GAS EXPIRY DATE: 02/04/2019  
VIOLATION CODE: DUI  
ARREST TIME: 20:59  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:25

SUBJECT'S LAST NAME: HANSEN

DATE/TIME OF TEST: 10/12/2016 21:57:33

---

Diagnostics Check	OK	21:50
Air Blank	0.000	21:51
Control Test	0.081	21:51
Air Blank	0.000	21:52
Subject Sample # 1	0.208	21:53
Breath Volume	1.667	21:53
Air Blank	0.000	21:53
Air Blank	0.000	21:55
Subject Sample # 2	0.208	21:55
Breath Volume	1.460	21:55
Air Blank	0.000	21:56
Control Test	0.081	21:56
Air Blank	0.000	21:57
Diagnostics Check	OK	21:57

---

Data Download Date/Time: 10/14/2016 16:04

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: O'NEAL,SHARIL  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: SCHAEFER,JOHN  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 09/09/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 279870  
DRY GAS EXPIRY DATE: 02/04/2019  
VIOLATION CODE: DUI  
ARREST TIME: 23:39  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:28

SUBJECT'S LAST NAME: CA

DATE/TIME OF TEST: 10/13/2016 01:01:46

---

Diagnostics Check	OK	00:54
Air Blank	0.000	00:55
Control Test	0.081	00:55
Air Blank	0.000	00:56
Subject Sample # 1	0.151	00:57
Breath Volume	2.234	00:57
Air Blank	0.000	00:57
Air Blank	0.000	00:59
Subject Sample # 2	0.158	01:00
Breath Volume	1.746	01:00
Air Blank	0.000	01:00
Control Test	0.079	01:01
Air Blank	0.000	01:01
Diagnostics Check	OK	01:01

---

Data Download Date/Time: 10/14/2016 16:04

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CAIN,JEANETTE,  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: SIEGEL,AARON  
ARRESTING OFFICER AGENCY: Delray Beach PD  
LAST AGENCY INSPECTION: 09/09/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 279870  
DRY GAS EXPIRY DATE: 02/04/2019  
VIOLATION CODE: DUI  
ARREST TIME: 16:03  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 16:35

SUBJECT'S LAST NAME: CAMERONE

DATE/TIME OF TEST: 10/13/2016 17:15:11

---

Diagnostics Check		OK	17:05
Air Blank		0.000	17:06
Control Test		0.081	17:06
Air Blank		0.000	17:06
Subject Sample	# 1	0.274	17:10
Breath Volume		1.222	17:10
Air Blank		0.000	17:10
Air Blank		0.000	17:12
Subject Sample	# 2	0.286	17:13
Breath Volume		2.101	17:13
Air Blank		0.000	17:14
Control Test		0.078	17:14
Air Blank		0.000	17:15
Diagnostics Check		OK	17:15

---

Data Download Date/Time: 10/14/2016 16:04



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CAIN,JEANETTE,  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: GEE,JOHN  
ARRESTING OFFICER AGENCY: Palm Springs PD  
LAST AGENCY INSPECTION: 09/09/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 279870  
DRY GAS EXPIRY DATE: 02/04/2019  
VIOLATION CODE: DUI  
ARREST TIME: 19:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:24

SUBJECT'S LAST NAME: FIENE

DATE/TIME OF TEST: 10/13/2016 20:10:40

---

Diagnostics Check	OK	20:06	
Air Blank	0.000	20:07	
Control Test	0.080	20:07	
Air Blank	0.000	20:08	
Subject Sample	# 1	20:09	Subject Test Refused
Breath Volume	0.000	20:09	
Air Blank	0.000	20:09	
Control Test	0.081	20:10	
Air Blank	0.000	20:10	
Diagnostics Check	OK	20:10	

---

Data Download Date/Time: 10/14/2016 16:04

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: O'NEAL,SHARIL  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: CROY,PAUL  
ARRESTING OFFICER AGENCY: FHP Troop L  
LAST AGENCY INSPECTION: 09/09/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 279870  
DRY GAS EXPIRY DATE: 02/04/2019  
VIOLATION CODE: DUI  
ARREST TIME: 01:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:37

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 10/14/2016 03:10:42

---

Diagnostics Check	OK	03:04
Air Blank	0.000	03:04
Control Test	0.081	03:05
Air Blank	0.000	03:05
Subject Sample # 1	0.170	03:06
Breath Volume	1.792	03:06
Air Blank	0.000	03:06
Air Blank	0.000	03:08
Subject Sample # 2	0.189	03:09
Breath Volume	1.558	03:09
Air Blank	0.000	03:09
Control Test	0.080	03:10
Air Blank	0.000	03:10
Diagnostics Check	OK	03:10

---

Data Download Date/Time: 10/14/2016 16:04

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006262

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BRYAN,STEVEN,A  
BREATH TEST OPERATOR AGENCY: Volusia County SO  
ARRESTING OFFICER: HILDEBRANDT,PATRICK  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 09/15/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 203498  
DRY GAS EXPIRY DATE: 12/10/2018  
VIOLATION CODE: DUI  
ARREST TIME: 18:03  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:16

SUBJECT'S LAST NAME: SALOMONE

DATE/TIME OF TEST: 10/13/2016 21:44:57

---

Diagnostics Check	OK	21:36
Air Blank	0.000	21:37
Control Test	0.080	21:37
Air Blank	0.000	21:38
Subject Sample # 1	0.198	21:39
Breath Volume	1.460	21:39
Air Blank	0.000	21:40
Air Blank	0.000	21:42
Subject Sample # 2	0.201	21:43
Breath Volume	1.339	21:43
Air Blank	0.000	21:44
Control Test	0.080	21:44
Air Blank	0.000	21:44
Diagnostics Check	OK	21:44

---

Data Download Date/Time: 10/19/2016 22:53

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006262

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DAVIS, J, J  
BREATH TEST OPERATOR AGENCY: Volusia County SO  
ARRESTING OFFICER: DAVIS, ANDREW  
ARRESTING OFFICER AGENCY: Volusia County SO  
LAST AGENCY INSPECTION: 09/15/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 203498  
DRY GAS EXPIRY DATE: 12/10/2018  
VIOLATION CODE: DUI  
ARREST TIME: 06:04  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 06:06

SUBJECT'S LAST NAME: WHITLOCK

DATE/TIME OF TEST: 10/16/2016 06:37:46

---

Diagnostics Check	OK	06:30
Air Blank	0.000	06:30
Control Test	0.081	06:31
Air Blank	0.000	06:31
Subject Sample # 1	0.099	06:32
Breath Volume	1.375	06:32
Air Blank	0.000	06:33
Air Blank	0.000	06:34
Subject Sample # 2	0.100	06:36
Breath Volume	1.683	06:36
Air Blank	0.000	06:36
Control Test	0.080	06:37
Air Blank	0.000	06:37
Diagnostics Check	OK	06:37

---

Data Download Date/Time: 10/19/2016 22:53

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006386

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CAMERON,DAVID,J  
BREATH TEST OPERATOR AGENCY: Hendry County SO  
ARRESTING OFFICER: KIRK,NATHAN  
ARRESTING OFFICER AGENCY: Hendry County SO  
LAST AGENCY INSPECTION: 09/28/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05115080A5  
DRY GAS EXPIRY DATE: 03/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 02:37  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: NEILLY

DATE/TIME OF TEST: 10/01/2016 03:33:07

---

Diagnostics Check	OK	03:26
Air Blank	0.000	03:26
Control Test	0.079	03:26
Air Blank	0.000	03:27
Subject Sample # 1	0.140	03:28
Breath Volume	2.187	03:28
Air Blank	0.000	03:29
Air Blank	0.000	03:30
Subject Sample # 2	0.127	03:31
Breath Volume	1.234	03:31
Air Blank	0.000	03:32
Control Test	0.077	03:32
Air Blank	0.000	03:33
Diagnostics Check	OK	03:33

---

Data Download Date/Time: 10/26/2016 14:41

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006386

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CAMERON,DAVID,J  
BREATH TEST OPERATOR AGENCY: Hendry County SO  
ARRESTING OFFICER: KIRK,NATHAN  
ARRESTING OFFICER AGENCY: Hendry County SO  
LAST AGENCY INSPECTION: 09/28/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05115080A5  
DRY GAS EXPIRY DATE: 03/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 22:41  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:51

SUBJECT'S LAST NAME: KENNINGTON

DATE/TIME OF TEST: 10/01/2016 23:14:41

---

Diagnostics Check	OK	23:11	
Air Blank	0.000	23:11	
Control Test	0.079	23:12	
Air Blank	0.000	23:12	
Subject Sample	# 1	23:13	Subject Test Refused
Breath Volume	0.000	23:13	
Air Blank	0.000	23:13	
Control Test	0.079	23:13	
Air Blank	0.000	23:14	
Diagnostics Check	OK	23:14	

---

Data Download Date/Time: 10/26/2016 14:41

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006386

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CAMERON,DAVID,J  
BREATH TEST OPERATOR AGENCY: Hendry County SO  
ARRESTING OFFICER: TAYLOR,ANDREW  
ARRESTING OFFICER AGENCY: Hendry County SO  
LAST AGENCY INSPECTION: 09/28/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05115080A5  
DRY GAS EXPIRY DATE: 03/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 22:53  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:07

SUBJECT'S LAST NAME: RAMIREZ-JIMENEZ

DATE/TIME OF TEST: 10/06/2016 23:35:38

---

Diagnostics Check	OK	23:32	
Air Blank	0.000	23:33	
Control Test	0.077	23:33	
Air Blank	0.000	23:33	
Subject Sample	# 1	23:34	Subject Test Refused
Breath Volume	0.000	23:34	
Air Blank	0.000	23:34	
Control Test	0.078	23:34	
Air Blank	0.000	23:35	
Diagnostics Check	OK	23:35	

---

Data Download Date/Time: 10/26/2016 14:41

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006386

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TAYLOR,M,R  
BREATH TEST OPERATOR AGENCY: Hendry County SO  
ARRESTING OFFICER: TAYLOR,A  
ARRESTING OFFICER AGENCY: Hendry County SO  
LAST AGENCY INSPECTION: 09/28/2016

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 05115080A5  
DRY GAS EXPIRY DATE: 03/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 03:22  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:48

SUBJECT'S LAST NAME: MENDOZA

DATE/TIME OF TEST: 10/08/2016 04:15:54

---

Diagnostics Check	OK	04:09
Air Blank	0.000	04:10
Control Test	0.078	04:10
Air Blank	0.000	04:10
Subject Sample # 1	0.015	04:11
Breath Volume	2.898	04:11
Air Blank	0.000	04:12
Air Blank	0.000	04:13
Subject Sample # 2	0.012	04:14
Breath Volume	3.503	04:14
Air Blank	0.000	04:14
Control Test	0.076	04:15
Air Blank	0.000	04:15
Diagnostics Check	OK	04:15

---

Data Download Date/Time: 10/26/2016 14:41



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006386

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CAMERON,DAVID,J  
BREATH TEST OPERATOR AGENCY: Hendry County SO  
ARRESTING OFFICER: CAMERON,DAVID  
ARRESTING OFFICER AGENCY: Hendry County SO  
LAST AGENCY INSPECTION: 09/28/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05115080A5  
DRY GAS EXPIRY DATE: 03/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 19:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:27

SUBJECT'S LAST NAME: CHANCE

DATE/TIME OF TEST: 10/15/2016 19:55:43

---

Diagnostics Check	OK	19:47
Air Blank	0.000	19:47
Control Test	0.077	19:48
Air Blank	0.000	19:48
Subject Sample # 1	0.151	19:51
Breath Volume	2.164	19:51
Air Blank	0.000	19:51
Air Blank	0.000	19:53
Subject Sample # 2	0.149	19:54
Breath Volume	2.183	19:54
Air Blank	0.000	19:54
Control Test	0.076	19:55
Air Blank	0.000	19:55
Diagnostics Check	OK	19:55

---

Data Download Date/Time: 10/26/2016 14:41

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006386

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CAMERON,DAVID,J  
BREATH TEST OPERATOR AGENCY: Hendry County SO  
ARRESTING OFFICER: CAMERON,DAVID  
ARRESTING OFFICER AGENCY: Hendry County SO  
LAST AGENCY INSPECTION: 09/28/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05115080A5  
DRY GAS EXPIRY DATE: 03/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 22:53  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:03

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 10/17/2016 00:29:47

---

Diagnostics Check	OK	00:23	
Air Blank	0.000	00:24	
Control Test	0.077	00:24	
Air Blank	0.000	00:25	
Subject Sample # 1	0.000	00:25	
Breath Volume	3.125	00:25	
Air Blank	0.000	00:26	
Air Blank	0.000	00:28	
Subject Sample # 2		00:28	Subject Test Refused
Breath Volume	0.000	00:28	
Air Blank	0.000	00:28	
Control Test	0.077	00:29	
Air Blank	0.000	00:29	
Diagnostics Check	OK	00:29	

---

Data Download Date/Time: 10/26/2016 14:41

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006386

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ELVER,GEORGE,T  
BREATH TEST OPERATOR AGENCY: Hendry County SO  
ARRESTING OFFICER: TYLER,CHRISTOPHER  
ARRESTING OFFICER AGENCY: Hendry County SO  
LAST AGENCY INSPECTION: 09/28/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05115080A5  
DRY GAS EXPIRY DATE: 03/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 21:51  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:04

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 10/21/2016 22:37:10

---

Diagnostics Check	OK	22:30
Air Blank	0.000	22:30
Control Test	0.078	22:31
Air Blank	0.000	22:31
Subject Sample # 1	0.165	22:32
Breath Volume	2.222	22:32
Air Blank	0.000	22:33
Air Blank	0.000	22:34
Subject Sample # 2	0.161	22:35
Breath Volume	2.273	22:35
Air Blank	0.000	22:36
Control Test	0.076	22:36
Air Blank	0.000	22:37
Diagnostics Check	OK	22:37

---

Data Download Date/Time: 10/26/2016 14:41

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PENDARVIS,TAMMANY,C  
BREATH TEST OPERATOR AGENCY: Escambia County SO  
ARRESTING OFFICER: HORN,JEREMY  
ARRESTING OFFICER AGENCY: Escambia County SO  
LAST AGENCY INSPECTION: 09/06/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05115080A5  
DRY GAS EXPIRY DATE: 03/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 04:22  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:41

SUBJECT'S LAST NAME: PENDARVIS

DATE/TIME OF TEST: 10/02/2016 05:11:13

---

Diagnostics Check	OK	05:08	
Air Blank	0.000	05:09	
Control Test	0.081	05:09	
Air Blank	0.000	05:09	
Subject Sample	# 1	05:10	Sequence Aborted
Breath Volume	0.000	05:10	
Air Blank	0.000	05:11	

---

Data Download Date/Time: 10/24/2016 07:09

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PENDARVIS,TAMMANY,C  
BREATH TEST OPERATOR AGENCY: Escambia County SO  
ARRESTING OFFICER: HORN,JEREMY  
ARRESTING OFFICER AGENCY: Escambia County SO  
LAST AGENCY INSPECTION: 09/06/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05115080A5  
DRY GAS EXPIRY DATE: 03/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 04:22  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:41

SUBJECT'S LAST NAME: PENDARVIS

DATE/TIME OF TEST: 10/02/2016 05:20:46

---

Diagnostics Check	OK	05:14
Air Blank	0.000	05:14
Control Test	0.081	05:15
Air Blank	0.000	05:15
Subject Sample # 1	0.059	05:16
Breath Volume	1.269	05:16
Air Blank	0.000	05:16
Air Blank	0.000	05:18
Subject Sample # 2	0.058	05:19
Breath Volume	1.292	05:19
Air Blank	0.000	05:19
Control Test	0.080	05:20
Air Blank	0.000	05:20
Diagnostics Check	OK	05:20

---

Data Download Date/Time: 10/24/2016 07:09

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SALON,DENNIS,M  
BREATH TEST OPERATOR AGENCY: Pensacola PD  
ARRESTING OFFICER: LAMBERT,ELIJAH  
ARRESTING OFFICER AGENCY: Pensacola PD  
LAST AGENCY INSPECTION: 09/06/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05115080A5  
DRY GAS EXPIRY DATE: 03/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 05:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:13

SUBJECT'S LAST NAME: RANDALL

DATE/TIME OF TEST: 10/11/2016 06:40:00

---

Diagnostics Check		OK	06:33
Air Blank		0.000	06:33
Control Test		0.080	06:34
Air Blank		0.000	06:34
Subject Sample	# 1	0.211	06:35
Breath Volume		1.710	06:35
Air Blank		0.000	06:36
Air Blank		0.000	06:37
Subject Sample	# 2	0.209	06:38
Breath Volume		1.957	06:38
Air Blank		0.000	06:39
Control Test		0.080	06:39
Air Blank		0.000	06:39
Diagnostics Check		OK	06:40

---

Data Download Date/Time: 10/24/2016 07:09

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: POOL,HAROLD,M  
BREATH TEST OPERATOR AGENCY: Escambia County SO  
ARRESTING OFFICER: POOL,HAROLD  
ARRESTING OFFICER AGENCY: Escambia County SO  
LAST AGENCY INSPECTION: 09/06/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05115080A5  
DRY GAS EXPIRY DATE: 03/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 22:21  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:06

SUBJECT'S LAST NAME: CARTER

DATE/TIME OF TEST: 10/14/2016 23:29:52

---

Diagnostics Check	OK	23:26	
Air Blank	0.000	23:27	
Control Test	0.081	23:27	
Air Blank	0.000	23:28	
Subject Sample	# 1	23:28	Subject Test Refused
Breath Volume	0.000	23:28	
Air Blank	0.000	23:28	
Control Test	0.081	23:29	
Air Blank	0.000	23:29	
Diagnostics Check	OK	23:29	

---

Data Download Date/Time: 10/24/2016 07:10

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROOP,CHAD,E  
BREATH TEST OPERATOR AGENCY: Escambia County SO  
ARRESTING OFFICER: ROOP,CHAD  
ARRESTING OFFICER AGENCY: Escambia County SO  
LAST AGENCY INSPECTION: 09/06/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05115080A5  
DRY GAS EXPIRY DATE: 03/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 02:56  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: OSWALD

DATE/TIME OF TEST: 10/15/2016 03:46:22

---

Diagnostics Check	OK	03:40
Air Blank	0.000	03:40
Control Test	0.081	03:41
Air Blank	0.000	03:41
Subject Sample # 1	0.000	03:41
Breath Volume	2.281	03:41
Air Blank	0.000	03:42
Air Blank	0.000	03:44
Subject Sample # 2	0.000	03:44
Breath Volume	1.878	03:44
Air Blank	0.000	03:45
Control Test	0.080	03:45
Air Blank	0.000	03:46
Diagnostics Check	OK	03:46

---

Data Download Date/Time: 10/24/2016 07:10



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SHELLEY,SAMUEL,L  
BREATH TEST OPERATOR AGENCY: Escambia County SO  
ARRESTING OFFICER: SHELLEY,SAMUEL  
ARRESTING OFFICER AGENCY: Escambia County SO  
LAST AGENCY INSPECTION: 10/24/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05115080A5  
DRY GAS EXPIRY DATE: 03/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 02:07  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: BOONE

DATE/TIME OF TEST: 10/27/2016 03:10:53

---

Diagnostics Check		OK	03:00
Air Blank		0.000	03:01
Control Test		0.081	03:01
Air Blank		0.000	03:02
Subject Sample	# 1	0.208	03:03
Breath Volume		3.613	03:03
Air Blank		0.000	03:03
Air Blank		0.000	03:05
Subject Sample	# 2	0.186	03:05
Breath Volume		1.933	03:05
Air Blank		0.000	03:06
Air Blank		0.000	03:08
Subject Sample	# 3	0.205	03:09
Breath Volume		3.589	03:09
Air Blank		0.000	03:09
Control Test		0.079	03:10
Air Blank		0.000	03:10
Diagnostics Check		OK	03:10

---

Data Download Date/Time: 11/22/2016 07:21

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: POOL,HAROLD,M  
BREATH TEST OPERATOR AGENCY: Escambia County SO  
ARRESTING OFFICER: POOL,HAROLD  
ARRESTING OFFICER AGENCY: Escambia County SO  
LAST AGENCY INSPECTION: 10/24/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05115080A5  
DRY GAS EXPIRY DATE: 03/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 01:42  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:37

SUBJECT'S LAST NAME: ROWE

DATE/TIME OF TEST: 10/29/2016 03:04:04

---

Diagnostics Check	OK	02:57
Air Blank	0.000	02:57
Control Test	0.081	02:58
Air Blank	0.000	02:58
Subject Sample # 1	0.049	02:59
Breath Volume	2.210	02:59
Air Blank	0.000	03:00
Air Blank	0.000	03:01
Subject Sample # 2	0.050	03:02
Breath Volume	3.382	03:02
Air Blank	0.000	03:03
Control Test	0.080	03:03
Air Blank	0.000	03:03
Diagnostics Check	OK	03:04

---

Data Download Date/Time: 11/22/2016 07:21

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HUDSON,JOSHUA,K  
BREATH TEST OPERATOR AGENCY: Pensacola PD  
ARRESTING OFFICER: GROGAN,STEVE  
ARRESTING OFFICER AGENCY: Pensacola PD  
LAST AGENCY INSPECTION: 10/24/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05115080A5  
DRY GAS EXPIRY DATE: 03/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 03:19  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: FARMER

DATE/TIME OF TEST: 10/30/2016 04:08:54

---

Diagnostics Check	OK	04:01
Air Blank	0.000	04:01
Control Test	0.081	04:02
Air Blank	0.000	04:02
Subject Sample # 1	0.137	04:03
Breath Volume	1.992	04:03
Air Blank	0.000	04:04
Air Blank	0.000	04:06
Subject Sample # 2	0.131	04:07
Breath Volume	2.128	04:07
Air Blank	0.000	04:07
Control Test	0.081	04:08
Air Blank	0.000	04:08
Diagnostics Check	OK	04:08

---

Data Download Date/Time: 11/22/2016 07:21

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DARLING,MEGHAN,P  
BREATH TEST OPERATOR AGENCY: Pensacola PD  
ARRESTING OFFICER: OSLEY,KAYLAN  
ARRESTING OFFICER AGENCY: Pensacola PD  
LAST AGENCY INSPECTION: 10/24/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05115080A5  
DRY GAS EXPIRY DATE: 03/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 05:46  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:31

SUBJECT'S LAST NAME: LANE

DATE/TIME OF TEST: 10/30/2016 07:06:18

---

Diagnostics Check		OK	06:57
Air Blank		0.000	06:58
Control Test		0.080	06:58
Air Blank		0.000	06:59
Subject Sample	# 1	0.213	07:01
Breath Volume		1.140	07:01
Air Blank		0.000	07:02
Air Blank		0.000	07:03
Subject Sample	# 2	0.194	07:04
Breath Volume		1.636	07:04
Air Blank		0.000	07:05
Control Test		0.076	07:05
Air Blank		0.000	07:06
Diagnostics Check		OK	07:06

---

Data Download Date/Time: 11/22/2016 07:21

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KENNEDY,JOSEPH,W  
BREATH TEST OPERATOR AGENCY: Miami PD  
ARRESTING OFFICER: KENNEDY,JOSEPH  
ARRESTING OFFICER AGENCY: Miami PD  
LAST AGENCY INSPECTION: 09/22/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 203498  
DRY GAS EXPIRY DATE: 12/10/2018  
VIOLATION CODE: DUI  
ARREST TIME: 14:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 14:39

SUBJECT'S LAST NAME: BARRERA ZALDIVAR

DATE/TIME OF TEST: 10/01/2016 15:03:06

---

Diagnostics Check	OK	14:59	
Air Blank	0.000	15:00	
Control Test	0.081	15:00	
Air Blank	0.000	15:01	
Subject Sample	# 1	15:01	Subject Test Refused
Breath Volume	0.000	15:01	
Air Blank	0.000	15:02	
Control Test	0.080	15:02	
Air Blank	0.000	15:03	
Diagnostics Check	OK	15:03	

---

Data Download Date/Time: 10/13/2016 16:50

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PEREZ,JOSHUA,  
BREATH TEST OPERATOR AGENCY: Miami PD  
ARRESTING OFFICER: PEREZ,JOSHUA  
ARRESTING OFFICER AGENCY: Miami PD  
LAST AGENCY INSPECTION: 09/22/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 203498  
DRY GAS EXPIRY DATE: 12/10/2018  
VIOLATION CODE: DUI  
ARREST TIME: 21:01  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: ALVAREZ

DATE/TIME OF TEST: 10/01/2016 21:58:29

---

Diagnostics Check	OK	21:52
Air Blank	0.000	21:52
Control Test	0.080	21:52
Air Blank	0.000	21:53
Subject Sample # 1	0.229	21:53
Breath Volume	2.027	21:53
Air Blank	0.000	21:54
Air Blank	0.000	21:56
Subject Sample # 2	0.229	21:56
Breath Volume	2.128	21:56
Air Blank	0.000	21:57
Control Test	0.078	21:57
Air Blank	0.000	21:58
Diagnostics Check	OK	21:58

---

Data Download Date/Time: 10/13/2016 16:50

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HERNANDEZ,ANGEL,A  
BREATH TEST OPERATOR AGENCY: Miami PD  
ARRESTING OFFICER: CABELLO,ADRIAN  
ARRESTING OFFICER AGENCY: Miami PD  
LAST AGENCY INSPECTION: 09/22/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 203498  
DRY GAS EXPIRY DATE: 12/10/2018  
VIOLATION CODE: DUI  
ARREST TIME: 04:24  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:45

SUBJECT'S LAST NAME: PERDOMO

DATE/TIME OF TEST: 10/02/2016 05:20:46

---

Diagnostics Check	OK	05:14
Air Blank	0.000	05:14
Control Test	0.081	05:14
Air Blank	0.000	05:15
Subject Sample # 1	0.181	05:16
Breath Volume	1.808	05:16
Air Blank	0.000	05:16
Air Blank	0.000	05:18
Subject Sample # 2	0.176	05:19
Breath Volume	1.878	05:19
Air Blank	0.000	05:19
Control Test	0.078	05:20
Air Blank	0.000	05:20
Diagnostics Check	OK	05:20

---

Data Download Date/Time: 10/13/2016 16:50

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PASTOR,JOSE,F  
BREATH TEST OPERATOR AGENCY: Miami PD  
ARRESTING OFFICER: PASTOR,JOSE  
ARRESTING OFFICER AGENCY: Miami PD  
LAST AGENCY INSPECTION: 09/22/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 203498  
DRY GAS EXPIRY DATE: 12/10/2018  
VIOLATION CODE: DUI  
ARREST TIME: 21:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:16

SUBJECT'S LAST NAME: VELASCO

DATE/TIME OF TEST: 10/02/2016 22:47:07

---

Diagnostics Check	OK	22:40
Air Blank	0.000	22:40
Control Test	0.081	22:41
Air Blank	0.000	22:41
Subject Sample # 1	0.344	22:42
Breath Volume	1.855	22:42
Air Blank	0.000	22:43
Air Blank	0.000	22:44
Subject Sample # 2	0.342	22:45
Breath Volume	2.113	22:45
Air Blank	0.000	22:46
Control Test	0.080	22:46
Air Blank	0.000	22:47
Diagnostics Check	OK	22:47

---

Data Download Date/Time: 10/13/2016 16:50





# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TARGIA,FRANK,  
BREATH TEST OPERATOR AGENCY: Miami PD  
ARRESTING OFFICER: ACUNA,JOSE  
ARRESTING OFFICER AGENCY: Miami PD  
LAST AGENCY INSPECTION: 09/22/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 203498  
DRY GAS EXPIRY DATE: 12/10/2018  
VIOLATION CODE: DUI  
ARREST TIME: 20:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: SOSA

DATE/TIME OF TEST: 10/06/2016 22:05:20

---

Diagnostics Check	OK	22:02	
Air Blank	0.000	22:02	
Control Test	0.079	22:02	
Air Blank	0.000	22:03	
Subject Sample	# 1	22:03	Subject Test Refused
Breath Volume	0.000	22:03	
Air Blank	0.000	22:04	
Control Test	0.078	22:04	
Air Blank	0.000	22:05	
Diagnostics Check	OK	22:05	

---

Data Download Date/Time: 10/13/2016 16:50

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HERNANDEZ,JASIEL,

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: HERNANDEZ,JASIEL

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 09/22/2016

Date or Time Changed: No

DRY GAS STANDARD LOT #: 203498

DRY GAS EXPIRY DATE: 12/10/2018

VIOLATION CODE: DUI

ARREST TIME: 23:38

GENDER: Male

OBSERVATION PERIOD BEGAN: 23:57

SUBJECT'S LAST NAME: MALDONADO

DATE/TIME OF TEST: 10/07/2016 00:25:13

---

Diagnostics Check		OK	00:18
Air Blank		0.000	00:18
Control Test		0.080	00:19
Air Blank		0.000	00:19
Subject Sample	# 1	0.208	00:20
Breath Volume		3.195	00:20
Air Blank		0.000	00:20
Air Blank		0.000	00:22
Subject Sample	# 2	0.201	00:23
Breath Volume		1.898	00:23
Air Blank		0.000	00:24
Control Test		0.079	00:24
Air Blank		0.000	00:25
Diagnostics Check		OK	00:25

---

Data Download Date/Time: 10/13/2016 16:50

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DIAZ,LEONARDO,A  
BREATH TEST OPERATOR AGENCY: Miami PD  
ARRESTING OFFICER: DIAZ,LEONARDO  
ARRESTING OFFICER AGENCY: Miami PD  
LAST AGENCY INSPECTION: 09/22/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 203498  
DRY GAS EXPIRY DATE: 12/10/2018  
VIOLATION CODE: DUI  
ARREST TIME: 10:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 10:30

SUBJECT'S LAST NAME: ROMAN

DATE/TIME OF TEST: 10/07/2016 11:06:51

---

Diagnostics Check	OK	11:00
Air Blank	0.000	11:00
Control Test	0.079	11:01
Air Blank	0.000	11:01
Subject Sample # 1	0.000	11:02
Breath Volume	2.800	11:02
Air Blank	0.000	11:03
Air Blank	0.000	11:05
Subject Sample # 2	0.000	11:05
Breath Volume	2.656	11:05
Air Blank	0.000	11:05
Control Test	0.079	11:06
Air Blank	0.000	11:06
Diagnostics Check	OK	11:06

---

Data Download Date/Time: 10/13/2016 16:50



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BAHAMONDE,JORDANY,

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: BAHAMONDE,JORDANY

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 09/22/2016

Date or Time Changed: No

DRY GAS STANDARD LOT #: 203498

DRY GAS EXPIRY DATE: 12/10/2018

VIOLATION CODE: DUI

ARREST TIME: 02:05

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: HERRERA

DATE/TIME OF TEST: 10/09/2016 03:21:48

---

Diagnostics Check	OK	03:15
Air Blank	0.000	03:16
Control Test	0.079	03:16
Air Blank	0.000	03:16
Subject Sample # 1	0.239	03:17
Breath Volume	1.148	03:17
Air Blank	0.000	03:18
Air Blank	0.000	03:19
Subject Sample # 2	0.242	03:20
Breath Volume	1.214	03:20
Air Blank	0.000	03:20
Control Test	0.079	03:21
Air Blank	0.000	03:21
Diagnostics Check	OK	03:21

---

Data Download Date/Time: 10/13/2016 16:50



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: CHERRELUS,RUDDENS,

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: CHERRELUS,RUDDENS

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 09/22/2016

Date or Time Changed: No

DRY GAS STANDARD LOT #: 203498

DRY GAS EXPIRY DATE: 12/10/2018

VIOLATION CODE: DUI

ARREST TIME: 03:19

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:36

SUBJECT'S LAST NAME: CORTEZ

DATE/TIME OF TEST: 10/10/2016 04:03:31

---

Diagnostics Check	OK	04:00	
Air Blank	0.000	04:01	
Control Test	0.080	04:01	
Air Blank	0.000	04:01	
Subject Sample	# 1	04:02	Subject Test Refused
Breath Volume	0.000	04:02	
Air Blank	0.000	04:02	
Control Test	0.079	04:02	
Air Blank	0.000	04:03	
Diagnostics Check	OK	04:03	

---

Data Download Date/Time: 10/13/2016 16:50



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LOPEZ,RYAN,  
BREATH TEST OPERATOR AGENCY: Miami PD  
ARRESTING OFFICER: HEREDIA-RUBIO,ROBERTO  
ARRESTING OFFICER AGENCY: Miami PD  
LAST AGENCY INSPECTION: 09/22/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 203498  
DRY GAS EXPIRY DATE: 12/10/2018  
VIOLATION CODE: DUI  
ARREST TIME: 19:41  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:14

SUBJECT'S LAST NAME: AGOSTINO

DATE/TIME OF TEST: 10/10/2016 20:40:09

---

Diagnostics Check	OK	20:34	
Air Blank	0.000	20:35	
Control Test	0.080	20:35	
Air Blank	0.000	20:35	
Subject Sample # 1	0.000	20:36	
Breath Volume	1.375	20:36	
Air Blank	0.000	20:36	
Air Blank	0.000	20:38	
		20:38	Improper Sample
Air Blank	0.000	20:39	
Control Test	0.080	20:39	
Air Blank	0.000	20:40	
Diagnostics Check	OK	20:40	

---

Data Download Date/Time: 10/13/2016 16:50

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LOPEZ,RYAN,  
BREATH TEST OPERATOR AGENCY: Miami PD  
ARRESTING OFFICER: HEREDIA-RUBIO,ROBERTO  
ARRESTING OFFICER AGENCY: Miami PD  
LAST AGENCY INSPECTION: 09/22/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 203498  
DRY GAS EXPIRY DATE: 12/10/2018  
VIOLATION CODE: DUI  
ARREST TIME: 19:41  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:14

SUBJECT'S LAST NAME: AGOSTINO

DATE/TIME OF TEST: 10/10/2016 20:49:45

---

Diagnostics Check	OK	20:43
Air Blank	0.000	20:44
Control Test	0.079	20:44
Air Blank	0.000	20:45
Subject Sample # 1	0.000	20:45
Breath Volume	1.207	20:45
Air Blank	0.000	20:46
Air Blank	0.000	20:48
Subject Sample # 2	0.000	20:48
Breath Volume	1.765	20:48
Air Blank	0.000	20:48
Control Test	0.078	20:49
Air Blank	0.000	20:49
Diagnostics Check	OK	20:49

---

Data Download Date/Time: 10/13/2016 16:50

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: AMORES,N,A  
BREATH TEST OPERATOR AGENCY: Miami PD  
ARRESTING OFFICER: AMORES,NESTOR  
ARRESTING OFFICER AGENCY: Miami PD  
LAST AGENCY INSPECTION: 09/22/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 203498  
DRY GAS EXPIRY DATE: 12/10/2018  
VIOLATION CODE: DUI  
ARREST TIME: 23:56  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:57

SUBJECT'S LAST NAME: MAZARIEGO

DATE/TIME OF TEST: 10/13/2016 00:24:03

---

Diagnostics Check	OK	00:17
Air Blank	0.000	00:17
Control Test	0.080	00:18
Air Blank	0.000	00:18
Subject Sample # 1	0.000	00:19
Breath Volume	1.304	00:19
Air Blank	0.000	00:20
Air Blank	0.000	00:22
Subject Sample # 2	0.000	00:22
Breath Volume	1.992	00:22
Air Blank	0.000	00:23
Control Test	0.079	00:23
Air Blank	0.000	00:23
Diagnostics Check	OK	00:24

---

Data Download Date/Time: 10/13/2016 16:50



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PALOMINO,JULIO,  
BREATH TEST OPERATOR AGENCY: Miami PD  
ARRESTING OFFICER: HERNANDEZ,JASIEL  
ARRESTING OFFICER AGENCY: Miami PD  
LAST AGENCY INSPECTION: 10/13/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 203498  
DRY GAS EXPIRY DATE: 12/10/2018  
VIOLATION CODE: DUI  
ARREST TIME: 05:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:04

SUBJECT'S LAST NAME: LOPEZ-VILLANUEVA

DATE/TIME OF TEST: 10/16/2016 06:37:16

---

Diagnostics Check	OK	06:24
Air Blank	0.000	06:25
Control Test	0.080	06:25
Air Blank	0.000	06:25
Subject Sample # 1	0.184	06:28
Breath Volume	2.589	06:28
Air Blank	0.000	06:29
Air Blank	0.000	06:31
Subject Sample # 2	0.157	06:32
Breath Volume	1.843	06:32
Air Blank	0.000	06:33
Air Blank	0.000	06:34
Subject Sample # 3	0.188	06:35
Breath Volume	2.796	06:35
Air Blank	0.000	06:36
Control Test	0.079	06:36
Air Blank	0.000	06:37
Diagnostics Check	OK	06:37

---

Data Download Date/Time: 11/29/2016 16:25

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HUSEL,BRIAN,M  
BREATH TEST OPERATOR AGENCY: Miami PD  
ARRESTING OFFICER: HUSEL,BRIAN  
ARRESTING OFFICER AGENCY: Miami PD  
LAST AGENCY INSPECTION: 10/13/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 203498  
DRY GAS EXPIRY DATE: 12/10/2018  
VIOLATION CODE: DUI  
ARREST TIME: 19:08  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 20:01

SUBJECT'S LAST NAME: BOWEN

DATE/TIME OF TEST: 10/16/2016 20:34:38

---

Diagnostics Check	OK	20:27
Air Blank	0.000	20:27
Control Test	0.081	20:28
Air Blank	0.000	20:28
Subject Sample # 1	0.000	20:30
Breath Volume	1.796	20:30
Air Blank	0.000	20:30
Air Blank	0.000	20:32
Subject Sample # 2	0.000	20:33
Breath Volume	1.261	20:33
Air Blank	0.000	20:33
Control Test	0.078	20:34
Air Blank	0.000	20:34
Diagnostics Check	OK	20:34

---

Data Download Date/Time: 11/29/2016 16:25

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HUSEL,BRIAN,M  
BREATH TEST OPERATOR AGENCY: Miami PD  
ARRESTING OFFICER: HUSEL,BRIAN  
ARRESTING OFFICER AGENCY: Miami PD  
LAST AGENCY INSPECTION: 10/13/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 203498  
DRY GAS EXPIRY DATE: 12/10/2018  
VIOLATION CODE: DUI  
ARREST TIME: 19:08  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 20:01

SUBJECT'S LAST NAME: BOWEN

DATE/TIME OF TEST: 10/16/2016 20:50:10

---

Diagnostics Check	OK	20:46	
Air Blank	0.000	20:47	
Control Test	0.080	20:47	
Air Blank	0.000	20:48	
Subject Sample	# 1	20:48	Subject Test Refused
Breath Volume	0.000	20:48	
Air Blank	0.000	20:49	
Control Test	0.080	20:49	
Air Blank	0.000	20:50	
Diagnostics Check	OK	20:50	

---

Data Download Date/Time: 11/29/2016 16:25

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: FERNANDEZ,ANGEL,

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: FERNANDEZ,ANGEL

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 10/13/2016

Date or Time Changed: No

DRY GAS STANDARD LOT #: 203498

DRY GAS EXPIRY DATE: 12/10/2018

VIOLATION CODE: DUI

ARREST TIME: 01:05

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:03

SUBJECT'S LAST NAME: IBANEZ

DATE/TIME OF TEST: 10/17/2016 02:28:02

---

Diagnostics Check	OK	02:24	
Air Blank	0.000	02:24	
Control Test	0.080	02:25	
Air Blank	0.000	02:25	
Subject Sample	# 1	02:26	Subject Test Refused
Breath Volume	0.000	02:26	
Air Blank	0.000	02:27	
Control Test	0.080	02:27	
Air Blank	0.000	02:27	
Diagnostics Check	OK	02:28	

---

Data Download Date/Time: 11/29/2016 16:25



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RODRIGUEZ,LUIS,G  
BREATH TEST OPERATOR AGENCY: Miami PD  
ARRESTING OFFICER: GUERRA,ALAIN  
ARRESTING OFFICER AGENCY: Miami PD  
LAST AGENCY INSPECTION: 10/13/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 203498  
DRY GAS EXPIRY DATE: 12/10/2018  
VIOLATION CODE: DUI  
ARREST TIME: 17:33  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:45

SUBJECT'S LAST NAME: SOTELO

DATE/TIME OF TEST: 10/20/2016 18:24:01

---

Diagnostics Check		OK	18:06	
Air Blank		0.000	18:07	
Control Test		0.079	18:07	
Air Blank		0.000	18:08	
Subject Sample	# 1	0.237	18:11	Volume Not Met
Breath Volume		0.195	18:11	
Air Blank		0.000	18:11	
Air Blank		0.000	18:13	
Subject Sample	# 2	0.238	18:16	
Breath Volume		1.207	18:16	
Air Blank		0.000	18:17	
Air Blank		0.000	18:19	
Subject Sample	# 3	0.254	18:22	Volume Not Met
Breath Volume		0.792	18:22	
Air Blank		0.000	18:23	
Control Test		0.077	18:23	
Air Blank		0.000	18:23	
Diagnostics Check		OK	18:24	

---

Data Download Date/Time: 11/29/2016 16:25

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RODRIGUEZ,LUIS,G  
BREATH TEST OPERATOR AGENCY: Miami PD  
ARRESTING OFFICER: GUERRA,ALAIN  
ARRESTING OFFICER AGENCY: Miami PD  
LAST AGENCY INSPECTION: 10/13/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 203498  
DRY GAS EXPIRY DATE: 12/10/2018  
VIOLATION CODE: DUI  
ARREST TIME: 17:33  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:45

SUBJECT'S LAST NAME: SOTELO

DATE/TIME OF TEST: 10/20/2016 18:43:37

---

Diagnostics Check		OK	18:28	
Air Blank		0.000	18:28	
Control Test		0.077	18:29	
Air Blank		0.000	18:29	
Subject Sample	# 1	0.251	18:31	No .020 agreement
Breath Volume		1.765	18:31	
Air Blank		0.000	18:31	
Air Blank		0.000	18:33	
Subject Sample	# 2		18:36	No Sample Provided
Breath Volume		0.000	18:36	
Air Blank		0.000	18:37	
Air Blank		0.000	18:39	
Subject Sample	# 3	0.228	18:41	No .020 agreement
Breath Volume		1.171	18:41	
Air Blank		0.000	18:42	
Control Test		0.079	18:42	
Air Blank		0.000	18:43	
Diagnostics Check		OK	18:43	

---

Data Download Date/Time: 11/29/2016 16:25

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HERNANDEZ,ANGEL,A  
BREATH TEST OPERATOR AGENCY: Miami PD  
ARRESTING OFFICER: HERNANDEZ,ANGEL  
ARRESTING OFFICER AGENCY: Miami PD  
LAST AGENCY INSPECTION: 10/13/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 203498  
DRY GAS EXPIRY DATE: 12/10/2018  
VIOLATION CODE: DUI  
ARREST TIME: 04:18  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: WILIAMS

DATE/TIME OF TEST: 10/21/2016 05:10:41

---

Diagnostics Check	OK	04:59	
Air Blank	0.000	04:59	
Control Test	0.079	05:00	
Air Blank	0.000	05:00	
Subject Sample	# 1	05:03	No Sample Provided
Breath Volume	0.000	05:03	
Air Blank	0.000	05:04	
Air Blank	0.000	05:06	
Subject Sample	# 2	05:09	No Sample Provided
Breath Volume	0.000	05:09	
Air Blank	0.000	05:09	
Control Test	0.079	05:10	
Air Blank	0.000	05:10	
Diagnostics Check	OK	05:10	

---

Data Download Date/Time: 11/29/2016 16:25

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HERNANDEZ,ANGEL,A  
BREATH TEST OPERATOR AGENCY: Miami PD  
ARRESTING OFFICER: HERNANDEZ,ANGEL  
ARRESTING OFFICER AGENCY: Miami PD  
LAST AGENCY INSPECTION: 10/13/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 203498  
DRY GAS EXPIRY DATE: 12/10/2018  
VIOLATION CODE: DUI  
ARREST TIME: 04:18  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 10/21/2016 05:16:21

---

Diagnostics Check	OK	05:13	
Air Blank	0.000	05:13	
Control Test	0.079	05:14	
Air Blank	0.000	05:14	
Subject Sample	# 1	05:14	Subject Test Refused
Breath Volume	0.000	05:14	
Air Blank	0.000	05:15	
Control Test	0.079	05:15	
Air Blank	0.000	05:16	
Diagnostics Check	OK	05:16	

---

Data Download Date/Time: 11/29/2016 16:25

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 203498
BREATH TEST OPERATOR: SUROS ENRIQUEZ,MYLEE,	DRY GAS EXPIRY DATE: 12/10/2018
BREATH TEST OPERATOR AGENCY: Miami PD	VIOLATION CODE: DUI
ARRESTING OFFICER: MARRERO,M	ARREST TIME: 06:59
ARRESTING OFFICER AGENCY: Miami PD	GENDER: Female
LAST AGENCY INSPECTION: 10/13/2016	OBSERVATION PERIOD BEGAN: 07:22

SUBJECT'S LAST NAME: SUROS ENRIQUEZ                      DATE/TIME OF TEST: 10/23/2016                      07:50:03

---

Diagnostics Check	OK	07:43
Air Blank	0.000	07:44
Control Test	0.080	07:44
Air Blank	0.000	07:44
Subject Sample                      # 1	0.130	07:45
Breath Volume	1.320	07:45
Air Blank	0.000	07:46
Air Blank	0.000	07:48
Subject Sample                      # 2	0.124	07:48
Breath Volume	1.421	07:48
Air Blank	0.000	07:49
Control Test	0.080	07:49
Air Blank	0.000	07:49
Diagnostics Check	OK	07:50

---

Data Download Date/Time: 11/29/2016 16:25

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: CANTERO,DANIEL,J

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: PICO,JAVIER

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 10/13/2016

Date or Time Changed: No

DRY GAS STANDARD LOT #: 203498

DRY GAS EXPIRY DATE: 12/10/2018

VIOLATION CODE: DUI

ARREST TIME: 18:54

GENDER: Male

OBSERVATION PERIOD BEGAN: 19:12

SUBJECT'S LAST NAME: GAMARRA

DATE/TIME OF TEST: 10/23/2016 19:34:57

---

Diagnostics Check	OK	19:32	
Air Blank	0.000	19:33	
Control Test	0.080	19:33	
Air Blank	0.000	19:33	
Subject Sample	# 1	19:34	RFI Detect
Breath Volume	0.000	19:34	
Air Blank		19:34	RFI Detect
Air Blank	0.000	19:34	

---

Data Download Date/Time: 11/29/2016 16:25

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: CANTERO,DANIEL,J

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: PICO,JAVIER

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 10/13/2016

Date or Time Changed: No

DRY GAS STANDARD LOT #: 203498

DRY GAS EXPIRY DATE: 12/10/2018

VIOLATION CODE: DUI

ARREST TIME: 18:54

GENDER: Male

OBSERVATION PERIOD BEGAN: 19:12

SUBJECT'S LAST NAME: GAMARRA

DATE/TIME OF TEST: 10/23/2016 19:39:40

---

Diagnostics Check	OK	19:36	
Air Blank	0.000	19:37	
Control Test	0.080	19:37	
Air Blank	0.000	19:38	
Subject Sample	# 1	19:38	Subject Test Refused
Breath Volume	0.000	19:38	
Air Blank	0.000	19:38	
Control Test	0.080	19:39	
Air Blank	0.000	19:39	
Diagnostics Check	OK	19:39	

---

Data Download Date/Time: 11/29/2016 16:25

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CHERRELUS,R,  
BREATH TEST OPERATOR AGENCY: Miami PD  
ARRESTING OFFICER: CHERRELUS,R  
ARRESTING OFFICER AGENCY: Miami PD  
LAST AGENCY INSPECTION: 10/13/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 203498  
DRY GAS EXPIRY DATE: 12/10/2018  
VIOLATION CODE: DUI  
ARREST TIME: 01:17  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:24

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 10/24/2016 01:50:05

---

Diagnostics Check	OK	01:47	
Air Blank	0.000	01:47	
Control Test	0.080	01:48	
Air Blank	0.000	01:48	
Subject Sample	# 1	01:48	Subject Test Refused
Breath Volume	0.000	01:48	
Air Blank	0.000	01:49	
Control Test	0.080	01:49	
Air Blank	0.000	01:50	
Diagnostics Check	OK	01:50	

---

Data Download Date/Time: 11/29/2016 16:25



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ACOSTA,CAROLINA,M  
BREATH TEST OPERATOR AGENCY: Miami PD  
ARRESTING OFFICER: URRUTIA,MARIO  
ARRESTING OFFICER AGENCY: Miami PD  
LAST AGENCY INSPECTION: 10/13/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 203498  
DRY GAS EXPIRY DATE: 12/10/2018  
VIOLATION CODE: DUI  
ARREST TIME: 20:25  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 20:40

SUBJECT'S LAST NAME: PADILLA

DATE/TIME OF TEST: 10/25/2016 21:07:38

---

Diagnostics Check	OK	21:01
Air Blank	0.000	21:01
Control Test	0.080	21:01
Air Blank	0.000	21:02
Subject Sample # 1	0.273	21:03
Breath Volume	2.039	21:03
Air Blank	0.000	21:03
Air Blank	0.000	21:05
Subject Sample # 2	0.292	21:06
Breath Volume	2.085	21:06
Air Blank	0.000	21:06
Control Test	0.080	21:07
Air Blank	0.000	21:07
Diagnostics Check	OK	21:07

---

Data Download Date/Time: 11/29/2016 16:25

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: QUINTANA,MICHAEL,A  
BREATH TEST OPERATOR AGENCY: Miami PD  
ARRESTING OFFICER: DUENA,DOUGLAS  
ARRESTING OFFICER AGENCY: Miami PD  
LAST AGENCY INSPECTION: 10/13/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 203498  
DRY GAS EXPIRY DATE: 12/10/2018  
VIOLATION CODE: DUI  
ARREST TIME: 05:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:15

SUBJECT'S LAST NAME: MORA-LEON

DATE/TIME OF TEST: 10/26/2016 07:05:20

---

Diagnostics Check	OK	06:58
Air Blank	0.000	06:58
Control Test	0.079	06:59
Air Blank	0.000	06:59
Subject Sample # 1	0.063	07:00
Breath Volume	1.347	07:00
Air Blank	0.000	07:00
Air Blank	0.000	07:02
Subject Sample # 2	0.068	07:03
Breath Volume	1.449	07:03
Air Blank	0.000	07:04
Control Test	0.079	07:04
Air Blank	0.000	07:05
Diagnostics Check	OK	07:05

---

Data Download Date/Time: 11/29/2016 16:25

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROQUE,EUGENE,H  
BREATH TEST OPERATOR AGENCY: Miami PD  
ARRESTING OFFICER: ROQUE,EUGENE  
ARRESTING OFFICER AGENCY: Miami PD  
LAST AGENCY INSPECTION: 10/13/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 203498  
DRY GAS EXPIRY DATE: 12/10/2018  
VIOLATION CODE: DUI  
ARREST TIME: 15:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:00

SUBJECT'S LAST NAME: FERIA

DATE/TIME OF TEST: 10/26/2016 16:27:51

---

Diagnostics Check	OK	16:21
Air Blank	0.000	16:21
Control Test	0.080	16:21
Air Blank	0.000	16:22
Subject Sample # 1	0.304	16:23
Breath Volume	1.425	16:23
Air Blank	0.000	16:24
Air Blank	0.000	16:25
Subject Sample # 2	0.307	16:26
Breath Volume	1.101	16:26
Air Blank	0.000	16:26
Control Test	0.076	16:27
Air Blank	0.000	16:27
Diagnostics Check	OK	16:27

---

Data Download Date/Time: 11/29/2016 16:25

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SIMINGTON,EARL,J  
BREATH TEST OPERATOR AGENCY: Miami PD  
ARRESTING OFFICER: SIMINGTON,EARL  
ARRESTING OFFICER AGENCY: Miami PD  
LAST AGENCY INSPECTION: 10/13/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 203498  
DRY GAS EXPIRY DATE: 12/10/2018  
VIOLATION CODE: DUI  
ARREST TIME: 23:54  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:21

SUBJECT'S LAST NAME: SALAZAR

DATE/TIME OF TEST: 10/29/2016 00:52:59

---

Diagnostics Check	OK	00:42	
Air Blank	0.000	00:42	
Control Test	0.079	00:42	
Air Blank	0.000	00:43	
Subject Sample	# 1	00:46	No Sample Provided
Breath Volume	0.000	00:46	
Air Blank	0.000	00:46	
Air Blank	0.000	00:48	
Subject Sample	# 2	00:51	Subject Test Refused
Breath Volume	0.000	00:51	
Air Blank	0.000	00:52	
Control Test	0.079	00:52	
Air Blank	0.000	00:52	
Diagnostics Check	OK	00:52	

---

Data Download Date/Time: 11/29/2016 16:25

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PALOMINO,JULIO,  
BREATH TEST OPERATOR AGENCY: Miami PD  
ARRESTING OFFICER: MONTESINO,MICHAEL  
ARRESTING OFFICER AGENCY: Miami PD  
LAST AGENCY INSPECTION: 10/13/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 203498  
DRY GAS EXPIRY DATE: 12/10/2018  
VIOLATION CODE: DUI  
ARREST TIME: 04:31  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:05

SUBJECT'S LAST NAME: ARCE

DATE/TIME OF TEST: 10/29/2016 05:30:03

---

Diagnostics Check	OK	05:26	
Air Blank	0.000	05:26	
Control Test	0.080	05:26	
Air Blank	0.000	05:27	
Subject Sample	# 1	05:28	Subject Test Refused
Breath Volume	0.000	05:28	
Air Blank	0.000	05:29	
Control Test	0.080	05:29	
Air Blank	0.000	05:29	
Diagnostics Check	OK	05:30	

---

Data Download Date/Time: 11/29/2016 16:25

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SEVILLA,MARIO,  
BREATH TEST OPERATOR AGENCY: Miami PD  
ARRESTING OFFICER: SEVILLA,MARIO  
ARRESTING OFFICER AGENCY: Miami PD  
LAST AGENCY INSPECTION: 10/13/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 203498  
DRY GAS EXPIRY DATE: 12/10/2018  
VIOLATION CODE: DUI  
ARREST TIME: 20:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:33

SUBJECT'S LAST NAME: NOGUERA

DATE/TIME OF TEST: 10/29/2016 21:30:09

---

Diagnostics Check	OK	21:27	
Air Blank	0.000	21:27	
Control Test	0.079	21:27	
Air Blank	0.000	21:28	
Subject Sample	# 1	21:28	Subject Test Refused
Breath Volume	0.000	21:28	
Air Blank	0.000	21:29	
Control Test	0.079	21:29	
Air Blank	0.000	21:30	
Diagnostics Check	OK	21:30	

---

Data Download Date/Time: 11/29/2016 16:25

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DELGADO,ANDY,  
BREATH TEST OPERATOR AGENCY: Miami PD  
ARRESTING OFFICER: MONTALVAN,CESAR  
ARRESTING OFFICER AGENCY: Miami PD  
LAST AGENCY INSPECTION: 10/13/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 203498  
DRY GAS EXPIRY DATE: 12/10/2018  
VIOLATION CODE: DUI  
ARREST TIME: 01:37  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:46

SUBJECT'S LAST NAME: ARANA

DATE/TIME OF TEST: 10/30/2016 02:16:18

---

Diagnostics Check	OK	02:08
Air Blank	0.000	02:09
Control Test	0.080	02:09
Air Blank	0.000	02:09
Subject Sample # 1	0.113	02:11
Breath Volume	2.023	02:11
Air Blank	0.000	02:12
Air Blank	0.000	02:13
Subject Sample # 2	0.108	02:14
Breath Volume	1.937	02:14
Air Blank	0.000	02:15
Control Test	0.079	02:15
Air Blank	0.000	02:16
Diagnostics Check	OK	02:16

---

Data Download Date/Time: 11/29/2016 16:25

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TARGIA,FRANK,  
BREATH TEST OPERATOR AGENCY: Miami PD  
ARRESTING OFFICER: VELADARES,CARLOS  
ARRESTING OFFICER AGENCY: Miami PD  
LAST AGENCY INSPECTION: 10/13/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 203498  
DRY GAS EXPIRY DATE: 12/10/2018  
VIOLATION CODE: DUI  
ARREST TIME: 01:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: MOLINA

DATE/TIME OF TEST: 10/30/2016 02:31:33

---

Diagnostics Check	OK	02:25	
Air Blank	0.000	02:26	
Control Test	0.079	02:26	
Air Blank	0.000	02:27	
Subject Sample	# 1	02:30	Subject Test Refused
Breath Volume	0.000	02:30	
Air Blank	0.000	02:30	
Control Test	0.079	02:30	
Air Blank	0.000	02:31	
Diagnostics Check	OK	02:31	

---

Data Download Date/Time: 11/29/2016 16:25



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BONILLA,ERASMO,G

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: MONTALVAN,CESAR

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 10/13/2016

Date or Time Changed: No

DRY GAS STANDARD LOT #: 203498

DRY GAS EXPIRY DATE: 12/10/2018

VIOLATION CODE: DUI

ARREST TIME: 01:44

GENDER: Male

OBSERVATION PERIOD BEGAN: 00:59

SUBJECT'S LAST NAME: ARRENDELL

DATE/TIME OF TEST: 10/31/2016 03:00:40

---

Diagnostics Check	OK	02:57	
Air Blank	0.000	02:58	
Control Test	0.080	02:58	
Air Blank	0.000	02:59	
Subject Sample	# 1	02:59	Subject Test Refused
Breath Volume	0.000	02:59	
Air Blank	0.000	02:59	
Control Test	0.079	03:00	
Air Blank	0.000	03:00	
Diagnostics Check	OK	03:00	

---

Data Download Date/Time: 11/29/2016 16:25

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LOPEZ,RYAN,  
BREATH TEST OPERATOR AGENCY: Miami PD  
ARRESTING OFFICER: HEREDIA RUBIO,ROBERTO  
ARRESTING OFFICER AGENCY: Miami PD  
LAST AGENCY INSPECTION: 09/22/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 203498  
DRY GAS EXPIRY DATE: 12/10/2018  
VIOLATION CODE: DUI  
ARREST TIME: 21:11  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:34

SUBJECT'S LAST NAME: CADENA

DATE/TIME OF TEST: 10/01/2016 21:58:01

---

Diagnostics Check	OK	21:55	
Air Blank	0.000	21:55	
Control Test	0.080	21:55	
Air Blank	0.000	21:56	
Subject Sample	# 1	21:56	Subject Test Refused
Breath Volume	0.000	21:56	
Air Blank	0.000	21:57	
Control Test	0.080	21:57	
Air Blank	0.000	21:57	
Diagnostics Check	OK	21:58	

---

Data Download Date/Time: 10/13/2016 17:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILLIAMS,BENNETT,W  
BREATH TEST OPERATOR AGENCY: Miami PD  
ARRESTING OFFICER: WILLIAMS,BENNETT  
ARRESTING OFFICER AGENCY: Miami PD  
LAST AGENCY INSPECTION: 09/22/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 203498  
DRY GAS EXPIRY DATE: 12/10/2018  
VIOLATION CODE: DUI  
ARREST TIME: 06:09  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:32

SUBJECT'S LAST NAME: BECERRA

DATE/TIME OF TEST: 10/02/2016 07:03:15

---

Diagnostics Check	OK	06:54	
Air Blank	0.000	06:55	
Control Test	0.080	06:55	
Air Blank	0.000	06:55	
Subject Sample	# 1	06:59	No Sample Provided
Breath Volume	0.000	06:59	
Air Blank	0.000	06:59	
Air Blank	0.000	07:01	
Subject Sample	# 2	07:01	Subject Test Refused
Breath Volume	0.000	07:01	
Air Blank	0.000	07:02	
Control Test	0.080	07:02	
Air Blank	0.000	07:03	
Diagnostics Check	OK	07:03	

---

Data Download Date/Time: 10/13/2016 17:37

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: VAZQUEZ,JONATHAN,L  
BREATH TEST OPERATOR AGENCY: Miami PD  
ARRESTING OFFICER: MUINA,RAFAEL  
ARRESTING OFFICER AGENCY: Miami PD  
LAST AGENCY INSPECTION: 09/22/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 203498  
DRY GAS EXPIRY DATE: 12/10/2018  
VIOLATION CODE: DUI  
ARREST TIME: 23:29  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:44

SUBJECT'S LAST NAME: MEJIAS

DATE/TIME OF TEST: 10/03/2016 00:14:33

---

Diagnostics Check	OK	00:06	
Air Blank	0.000	00:07	
Control Test	0.080	00:07	
Air Blank	0.000	00:08	
Subject Sample	# 1 0.294	00:08	
Breath Volume	2.390	00:08	
Air Blank	0.000	00:09	
Air Blank	0.000	00:10	
Subject Sample	# 2	00:12	Subject Test Refused
Breath Volume	0.031	00:12	
Air Blank	0.000	00:13	
Control Test	0.080	00:13	
Air Blank	0.000	00:14	
Diagnostics Check	OK	00:14	

---

Data Download Date/Time: 10/13/2016 17:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GIL,JORGE,  
BREATH TEST OPERATOR AGENCY: Miami PD  
ARRESTING OFFICER: GIL,JORGE  
ARRESTING OFFICER AGENCY: Miami PD  
LAST AGENCY INSPECTION: 09/22/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 203498  
DRY GAS EXPIRY DATE: 12/10/2018  
VIOLATION CODE: DUI  
ARREST TIME: 04:10  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: SANTOS

DATE/TIME OF TEST: 10/09/2016 04:43:49

---

Diagnostics Check	OK	04:36
Air Blank	0.000	04:36
Control Test	0.080	04:36
Air Blank	0.000	04:37
Subject Sample # 1	0.186	04:39
Breath Volume	1.531	04:39
Air Blank	0.000	04:39
Air Blank	0.000	04:41
Subject Sample # 2	0.188	04:42
Breath Volume	1.597	04:42
Air Blank	0.000	04:42
Control Test	0.078	04:43
Air Blank	0.000	04:43
Diagnostics Check	OK	04:43

---

Data Download Date/Time: 10/13/2016 17:37

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARNANDAN,INDROWTIE,  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: JESSUP,BRIAN  
ARRESTING OFFICER AGENCY: Davie PD  
LAST AGENCY INSPECTION: 09/24/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 34015080A3  
DRY GAS EXPIRY DATE: 01/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 00:59  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:27

SUBJECT'S LAST NAME: SOUCIE

DATE/TIME OF TEST: 10/01/2016 01:54:52

---

Diagnostics Check	OK	01:47
Air Blank	0.000	01:48
Control Test	0.079	01:48
Air Blank	0.000	01:48
Subject Sample # 1	0.115	01:50
Breath Volume	1.398	01:50
Air Blank	0.000	01:50
Air Blank	0.000	01:52
Subject Sample # 2	0.113	01:53
Breath Volume	1.148	01:53
Air Blank	0.000	01:53
Control Test	0.077	01:54
Air Blank	0.000	01:54
Diagnostics Check	OK	01:54

---

Data Download Date/Time: 10/19/2016 11:34

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GRAY,SANDRA,  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: LUMPKINS,HENRY  
ARRESTING OFFICER AGENCY: Hollywood PD  
LAST AGENCY INSPECTION: 09/24/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 34015080A3  
DRY GAS EXPIRY DATE: 01/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 02:54  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: SADAKA

DATE/TIME OF TEST: 10/01/2016 03:46:56

---

Diagnostics Check	OK	03:40
Air Blank	0.000	03:41
Control Test	0.079	03:41
Air Blank	0.000	03:42
Subject Sample # 1	0.000	03:42
Breath Volume	1.726	03:42
Air Blank	0.000	03:43
Air Blank	0.000	03:45
Subject Sample # 2	0.000	03:45
Breath Volume	2.402	03:45
Air Blank	0.000	03:46
Control Test	0.079	03:46
Air Blank	0.000	03:46
Diagnostics Check	OK	03:46

---

Data Download Date/Time: 10/19/2016 11:34

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MEDINA,MARMILLY,  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: SERPHOS,STEVEN  
ARRESTING OFFICER AGENCY: Broward County SO  
LAST AGENCY INSPECTION: 09/24/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 34015080A3  
DRY GAS EXPIRY DATE: 01/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 00:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 10/02/2016 01:09:21

---

Diagnostics Check	OK	01:05	
Air Blank	0.000	01:05	
Control Test	0.078	01:05	
Air Blank	0.000	01:06	
Subject Sample	# 1	01:07	Subject Test Refused
Breath Volume	0.132	01:07	
Air Blank	0.000	01:08	
Control Test	0.078	01:08	
Air Blank	0.000	01:09	
Diagnostics Check	OK	01:09	

---

Data Download Date/Time: 10/19/2016 11:34



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MEDINA,MARMILLY,M  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: SERPHOS,STEVEN  
ARRESTING OFFICER AGENCY: Broward County SO  
LAST AGENCY INSPECTION: 09/24/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 34015080A3  
DRY GAS EXPIRY DATE: 01/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 00:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 10/02/2016 01:37:31

---

Diagnostics Check	OK	01:30
Air Blank	0.000	01:31
Control Test	0.078	01:31
Air Blank	0.000	01:31
Subject Sample # 1	0.278	01:32
Breath Volume	1.558	01:32
Air Blank	0.000	01:33
Air Blank	0.000	01:35
Subject Sample # 2	0.283	01:35
Breath Volume	2.375	01:35
Air Blank	0.000	01:36
Control Test	0.076	01:36
Air Blank	0.000	01:37
Diagnostics Check	OK	01:37

---

Data Download Date/Time: 10/19/2016 11:34

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TILLMAN,SHEILA,L  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: WILSON,LAMAR  
ARRESTING OFFICER AGENCY: OTHER State  
LAST AGENCY INSPECTION: 09/24/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 34015080A3  
DRY GAS EXPIRY DATE: 01/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 02:08  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: ARNOLD

DATE/TIME OF TEST: 10/02/2016 03:10:54

---

Diagnostics Check		OK	03:03
Air Blank		0.000	03:04
Control Test		0.079	03:04
Air Blank		0.000	03:04
Subject Sample	# 1	0.178	03:05
Breath Volume		1.476	03:05
Air Blank		0.000	03:06
Air Blank		0.000	03:08
Subject Sample	# 2	0.166	03:09
Breath Volume		1.585	03:09
Air Blank		0.000	03:09
Control Test		0.079	03:10
Air Blank		0.000	03:10
Diagnostics Check		OK	03:10

---

Data Download Date/Time: 10/19/2016 11:34

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FRAZIER,ANAYA,S  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: SERPHOS,STEVEN  
ARRESTING OFFICER AGENCY: Broward County SO  
LAST AGENCY INSPECTION: 09/24/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 34015080A3  
DRY GAS EXPIRY DATE: 01/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 03:48  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:17

SUBJECT'S LAST NAME: EARLY

DATE/TIME OF TEST: 10/02/2016 04:46:34

---

Diagnostics Check	OK	04:37
Air Blank	0.000	04:38
Control Test	0.079	04:38
Air Blank	0.000	04:39
Subject Sample # 1	0.157	04:39
Breath Volume	1.667	04:39
Air Blank	0.000	04:40
Air Blank	0.000	04:41
Subject Sample # 2	0.144	04:44
Breath Volume	1.453	04:44
Air Blank	0.000	04:45
Control Test	0.079	04:45
Air Blank	0.000	04:46
Diagnostics Check	OK	04:46

---

Data Download Date/Time: 10/19/2016 11:34

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FRAZIER,ANAYA,S  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: LUMPKINS,HENRY  
ARRESTING OFFICER AGENCY: Hollywood PD  
LAST AGENCY INSPECTION: 09/24/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 34015080A3  
DRY GAS EXPIRY DATE: 01/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 01:19  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:51

SUBJECT'S LAST NAME: SCOTT

DATE/TIME OF TEST: 10/03/2016 02:26:49

---

Diagnostics Check	OK	02:16
Air Blank	0.000	02:16
Control Test	0.078	02:17
Air Blank	0.000	02:17
Subject Sample # 1	0.112	02:19
Breath Volume	1.199	02:19
Air Blank	0.000	02:20
Air Blank	0.000	02:22
Subject Sample # 2	0.116	02:25
Breath Volume	1.335	02:25
Air Blank	0.000	02:25
Control Test	0.079	02:26
Air Blank	0.000	02:26
Diagnostics Check	OK	02:26

---

Data Download Date/Time: 10/19/2016 11:34

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TILLMAN,SHEILA,L  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: MELENDEZ,MANNY  
ARRESTING OFFICER AGENCY: OTHER State  
LAST AGENCY INSPECTION: 09/24/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 34015080A3  
DRY GAS EXPIRY DATE: 01/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 00:05  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: MINO

DATE/TIME OF TEST: 10/08/2016 01:10:13

---

Diagnostics Check	OK	01:03
Air Blank	0.000	01:03
Control Test	0.078	01:04
Air Blank	0.000	01:04
Subject Sample # 1	0.187	01:05
Breath Volume	1.707	01:05
Air Blank	0.000	01:05
Air Blank	0.000	01:07
Subject Sample # 2	0.178	01:08
Breath Volume	1.566	01:08
Air Blank	0.000	01:09
Control Test	0.078	01:09
Air Blank	0.000	01:10
Diagnostics Check	OK	01:10

---

Data Download Date/Time: 10/19/2016 11:34

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARNANDAN,INDROWTIE,  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: WILLIAMS,SEAN  
ARRESTING OFFICER AGENCY: Broward County SO  
LAST AGENCY INSPECTION: 09/24/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 34015080A3  
DRY GAS EXPIRY DATE: 01/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 03:51  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: ADORNO

DATE/TIME OF TEST: 10/08/2016 04:56:41

---

Diagnostics Check	OK	04:50
Air Blank	0.000	04:50
Control Test	0.078	04:51
Air Blank	0.000	04:51
Subject Sample # 1	0.151	04:51
Breath Volume	1.558	04:51
Air Blank	0.000	04:52
Air Blank	0.000	04:54
Subject Sample # 2	0.140	04:55
Breath Volume	1.523	04:55
Air Blank	0.000	04:55
Control Test	0.077	04:56
Air Blank	0.000	04:56
Diagnostics Check	OK	04:56

---

Data Download Date/Time: 10/19/2016 11:34

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FRAZIER,ANAYA,S  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: PRESUME,JEHUD  
ARRESTING OFFICER AGENCY: Miramar PD  
LAST AGENCY INSPECTION: 09/24/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 34015080A3  
DRY GAS EXPIRY DATE: 01/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 02:16  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:01

SUBJECT'S LAST NAME: OBANDO

DATE/TIME OF TEST: 10/10/2016 03:34:07

---

Diagnostics Check	OK	03:27
Air Blank	0.000	03:28
Control Test	0.078	03:28
Air Blank	0.000	03:28
Subject Sample # 1	0.212	03:29
Breath Volume	1.500	03:29
Air Blank	0.000	03:30
Air Blank	0.000	03:32
Subject Sample # 2	0.210	03:32
Breath Volume	1.230	03:32
Air Blank	0.000	03:33
Control Test	0.078	03:33
Air Blank	0.000	03:34
Diagnostics Check	OK	03:34

---

Data Download Date/Time: 10/19/2016 11:34

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MEDINA,MARMILLY,  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: MCGILLICUDDY,STEVEN  
ARRESTING OFFICER AGENCY: Miramar PD  
LAST AGENCY INSPECTION: 09/24/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 34015080A3  
DRY GAS EXPIRY DATE: 01/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 03:25  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: MCKENZIE

DATE/TIME OF TEST: 10/11/2016 04:48:50

---

Diagnostics Check	OK	04:41
Air Blank	0.000	04:41
Control Test	0.078	04:41
Air Blank	0.000	04:42
Subject Sample # 1	0.201	04:43
Breath Volume	1.371	04:43
Air Blank	0.000	04:44
Air Blank	0.000	04:46
Subject Sample # 2	0.186	04:47
Breath Volume	1.230	04:47
Air Blank	0.000	04:47
Control Test	0.077	04:48
Air Blank	0.000	04:48
Diagnostics Check	OK	04:48

---

Data Download Date/Time: 10/19/2016 11:34



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MEDINA,MARMILLY,  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: PRESUME,JEHUD  
ARRESTING OFFICER AGENCY: Miramar PD  
LAST AGENCY INSPECTION: 09/24/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 34015080A3  
DRY GAS EXPIRY DATE: 01/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 22:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: RUJANO

DATE/TIME OF TEST: 10/11/2016 23:53:50

---

Diagnostics Check	OK	23:47
Air Blank	0.000	23:47
Control Test	0.078	23:47
Air Blank	0.000	23:48
Subject Sample # 1	0.135	23:49
Breath Volume	2.187	23:49
Air Blank	0.000	23:49
Air Blank	0.000	23:51
Subject Sample # 2	0.135	23:52
Breath Volume	2.039	23:52
Air Blank	0.000	23:52
Control Test	0.077	23:53
Air Blank	0.000	23:53
Diagnostics Check	OK	23:53

---

Data Download Date/Time: 10/19/2016 11:34

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TILLMAN,SHEILA,L  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: WASHINGTON,YVETTE  
ARRESTING OFFICER AGENCY: Broward County SO  
LAST AGENCY INSPECTION: 09/24/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 34015080A3  
DRY GAS EXPIRY DATE: 01/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 03:08  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 10/13/2016 04:07:47

---

Diagnostics Check	OK	04:00
Air Blank	0.000	04:01
Control Test	0.079	04:01
Air Blank	0.000	04:02
Subject Sample # 1	0.054	04:02
Breath Volume	1.324	04:02
Air Blank	0.000	04:03
Air Blank	0.000	04:04
Subject Sample # 2	0.052	04:06
Breath Volume	1.253	04:06
Air Blank	0.000	04:06
Control Test	0.079	04:07
Air Blank	0.000	04:07
Diagnostics Check	OK	04:07

---

Data Download Date/Time: 10/19/2016 11:34

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FRAZIER,ANAYA,S  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: JESSUP,BRIAN  
ARRESTING OFFICER AGENCY: Davie PD  
LAST AGENCY INSPECTION: 09/24/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 34015080A3  
DRY GAS EXPIRY DATE: 01/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 22:31  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: COX V

DATE/TIME OF TEST: 10/14/2016 23:42:38

---

Diagnostics Check	OK	23:35
Air Blank	0.000	23:36
Control Test	0.079	23:36
Air Blank	0.000	23:37
Subject Sample # 1	0.000	23:37
Breath Volume	2.707	23:37
Air Blank	0.000	23:38
Air Blank	0.000	23:40
Subject Sample # 2	0.000	23:41
Breath Volume	1.933	23:41
Air Blank	0.000	23:41
Control Test	0.078	23:42
Air Blank	0.000	23:42
Diagnostics Check	OK	23:42

---

Data Download Date/Time: 10/19/2016 11:34

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FRAZIER,ANAYA,S  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: PERCIVAL,KAITLYN  
ARRESTING OFFICER AGENCY: Pembroke Pines PD  
LAST AGENCY INSPECTION: 09/24/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 34015080A3  
DRY GAS EXPIRY DATE: 01/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 04:11  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:58

SUBJECT'S LAST NAME: MCDONALD

DATE/TIME OF TEST: 10/16/2016 05:26:16

---

Diagnostics Check	OK	05:19
Air Blank	0.000	05:19
Control Test	0.078	05:19
Air Blank	0.000	05:20
Subject Sample # 1	0.166	05:21
Breath Volume	1.757	05:21
Air Blank	0.000	05:21
Air Blank	0.000	05:23
Subject Sample # 2	0.183	05:24
Breath Volume	1.351	05:24
Air Blank	0.000	05:25
Control Test	0.078	05:25
Air Blank	0.000	05:26
Diagnostics Check	OK	05:26

---

Data Download Date/Time: 10/19/2016 11:34

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TILLMAN,SHEILA,L  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: WOOLLEY,KYLE  
ARRESTING OFFICER AGENCY: Margate PD  
LAST AGENCY INSPECTION: 09/24/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 34015080A3  
DRY GAS EXPIRY DATE: 01/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 17:03  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:00

SUBJECT'S LAST NAME: MARRERO

DATE/TIME OF TEST: 10/16/2016 18:40:53

---

Diagnostics Check	OK	18:32
Air Blank	0.000	18:33
Control Test	0.079	18:33
Air Blank	0.000	18:34
Subject Sample # 1	0.144	18:35
Breath Volume	1.429	18:35
Air Blank	0.000	18:35
Air Blank	0.000	18:37
Subject Sample # 2	0.147	18:39
Breath Volume	1.218	18:39
Air Blank	0.000	18:39
Control Test	0.080	18:40
Air Blank	0.000	18:40
Diagnostics Check	OK	18:40

---

Data Download Date/Time: 10/19/2016 11:34

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BIGGS,JAMES,G  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: CHIQUITO-  
RODRIGUEZ,LENNDY  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 10/14/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 279870  
DRY GAS EXPIRY DATE: 02/04/2019  
VIOLATION CODE: DUI  
ARREST TIME: 19:29  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 20:15

SUBJECT'S LAST NAME: MURPHY

DATE/TIME OF TEST: 10/14/2016 20:47:19

---

Diagnostics Check	OK	20:40
Air Blank	0.000	20:41
Control Test	0.081	20:41
Air Blank	0.000	20:41
Subject Sample # 1	0.177	20:42
Breath Volume	3.058	20:42
Air Blank	0.000	20:43
Air Blank	0.000	20:44
Subject Sample # 2	0.175	20:45
Breath Volume	3.007	20:45
Air Blank	0.000	20:46
Control Test	0.080	20:46
Air Blank	0.000	20:47
Diagnostics Check	OK	20:47

---

Data Download Date/Time: 11/18/2016 12:30

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BIGGS,JAMES,G  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: MCMAHON,GEORGE  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 10/14/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 279870  
DRY GAS EXPIRY DATE: 02/04/2019  
VIOLATION CODE: DUI  
ARREST TIME: 20:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: MASIARCZYK

DATE/TIME OF TEST: 10/14/2016 21:20:30

---

Diagnostics Check	OK	21:13
Air Blank	0.000	21:14
Control Test	0.080	21:14
Air Blank	0.000	21:15
Subject Sample # 1	0.132	21:15
Breath Volume	2.492	21:15
Air Blank	0.000	21:16
Air Blank	0.000	21:18
Subject Sample # 2	0.131	21:18
Breath Volume	2.050	21:18
Air Blank	0.000	21:19
Control Test	0.079	21:19
Air Blank	0.000	21:20
Diagnostics Check	OK	21:20

---

Data Download Date/Time: 11/18/2016 12:30

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BIGGS,JAMES,G  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: CROY,PAUL  
ARRESTING OFFICER AGENCY: FHP Troop L  
LAST AGENCY INSPECTION: 10/14/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 279870  
DRY GAS EXPIRY DATE: 02/04/2019  
VIOLATION CODE: DUI  
ARREST TIME: 20:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:11

SUBJECT'S LAST NAME: RUTLEDGE

DATE/TIME OF TEST: 10/14/2016 21:53:41

---

Diagnostics Check	OK	21:47
Air Blank	0.000	21:47
Control Test	0.079	21:47
Air Blank	0.000	21:48
Subject Sample # 1	0.199	21:48
Breath Volume	2.582	21:48
Air Blank	0.000	21:49
Air Blank	0.000	21:51
Subject Sample # 2	0.203	21:51
Breath Volume	3.152	21:51
Air Blank	0.000	21:52
Control Test	0.079	21:53
Air Blank	0.000	21:53
Diagnostics Check	OK	21:53

---

Data Download Date/Time: 11/18/2016 12:30



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: OWEN,SUE,  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: WRIGHT,DELSHAWNTA  
ARRESTING OFFICER AGENCY: FHP Troop L  
LAST AGENCY INSPECTION: 10/14/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 279870  
DRY GAS EXPIRY DATE: 02/04/2019  
VIOLATION CODE: DUI  
ARREST TIME: 06:43  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 07:05

SUBJECT'S LAST NAME: PEREZ DE LEON

DATE/TIME OF TEST: 10/15/2016 07:37:54

---

Diagnostics Check	OK	07:30
Air Blank	0.000	07:31
Control Test	0.081	07:31
Air Blank	0.000	07:31
Subject Sample # 1	0.143	07:33
Breath Volume	1.722	07:33
Air Blank	0.000	07:33
Air Blank	0.000	07:35
Subject Sample # 2	0.150	07:36
Breath Volume	2.000	07:36
Air Blank	0.000	07:36
Control Test	0.080	07:37
Air Blank	0.000	07:37
Diagnostics Check	OK	07:37

---

Data Download Date/Time: 11/18/2016 12:30

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PARENT,GARY,J  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: LASSO,JAVIER  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 10/14/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 279870  
DRY GAS EXPIRY DATE: 02/04/2019  
VIOLATION CODE: DUI  
ARREST TIME: 19:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:10

SUBJECT'S LAST NAME: MENDEZ PEREZ

DATE/TIME OF TEST: 10/16/2016 21:12:22

---

Diagnostics Check	OK	21:05
Air Blank	0.000	21:05
Control Test	0.081	21:06
Air Blank	0.000	21:06
Subject Sample # 1	0.222	21:07
Breath Volume	3.773	21:07
Air Blank	0.000	21:08
Air Blank	0.000	21:09
Subject Sample # 2	0.220	21:10
Breath Volume	3.574	21:10
Air Blank	0.000	21:11
Control Test	0.079	21:11
Air Blank	0.000	21:12
Diagnostics Check	OK	21:12

---

Data Download Date/Time: 11/18/2016 12:30

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: OWEN,SUE,  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: SENTMANAT,ADOLFO  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 10/14/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 279870  
DRY GAS EXPIRY DATE: 02/04/2019  
VIOLATION CODE: DUI  
ARREST TIME: 21:32  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:01

SUBJECT'S LAST NAME: RIVERA-GONZALEZ

DATE/TIME OF TEST: 10/16/2016 22:35:00

---

Diagnostics Check	OK	22:27
Air Blank	0.000	22:27
Control Test	0.081	22:28
Air Blank	0.000	22:28
Subject Sample # 1	0.282	22:30
Breath Volume	1.964	22:30
Air Blank	0.000	22:31
Air Blank	0.000	22:32
Subject Sample # 2	0.279	22:33
Breath Volume	1.406	22:33
Air Blank	0.000	22:34
Control Test	0.078	22:34
Air Blank	0.000	22:34
Diagnostics Check	OK	22:35

---

Data Download Date/Time: 11/18/2016 12:30

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: OWEN,SUE,  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: LASSO,JAVIER  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 10/14/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 279870  
DRY GAS EXPIRY DATE: 02/04/2019  
VIOLATION CODE: DUI  
ARREST TIME: 00:26  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:34

SUBJECT'S LAST NAME: MENDEZ GARCIA

DATE/TIME OF TEST: 10/17/2016 02:06:22

---

Diagnostics Check	OK	01:58
Air Blank	0.000	01:59
Control Test	0.081	01:59
Air Blank	0.000	02:00
Subject Sample # 1	0.193	02:00
Breath Volume	1.878	02:00
Air Blank	0.000	02:01
Air Blank	0.000	02:03
Subject Sample # 2	0.200	02:04
Breath Volume	2.371	02:04
Air Blank	0.000	02:05
Control Test	0.080	02:05
Air Blank	0.000	02:06
Diagnostics Check	OK	02:06

---

Data Download Date/Time: 11/18/2016 12:30

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: OWEN,SUE,  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: VANCAMP,JEFFERY  
ARRESTING OFFICER AGENCY: Boca Raton PD  
LAST AGENCY INSPECTION: 10/14/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 279870  
DRY GAS EXPIRY DATE: 02/04/2019  
VIOLATION CODE: DUI  
ARREST TIME: 01:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: HERNANDEZ SANCHEZ

DATE/TIME OF TEST: 10/17/2016 04:12:52

---

Diagnostics Check	OK	04:05
Air Blank	0.000	04:06
Control Test	0.080	04:06
Air Blank	0.000	04:06
Subject Sample # 1	0.203	04:07
Breath Volume	1.343	04:07
Air Blank	0.000	04:08
Air Blank	0.000	04:10
Subject Sample # 2	0.195	04:11
Breath Volume	1.621	04:11
Air Blank	0.000	04:11
Control Test	0.079	04:12
Air Blank	0.000	04:12
Diagnostics Check	OK	04:12

---

Data Download Date/Time: 11/18/2016 12:30

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: OWEN,SUE,  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: WINTERS,JEFFERY  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 10/14/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 279870  
DRY GAS EXPIRY DATE: 02/04/2019  
VIOLATION CODE: DUI  
ARREST TIME: 13:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 14:30

SUBJECT'S LAST NAME: JARVIS

DATE/TIME OF TEST: 10/17/2016 15:10:47

---

Diagnostics Check	OK	15:04
Air Blank	0.000	15:04
Control Test	0.082	15:05
Air Blank	0.000	15:05
Subject Sample # 1	0.000	15:06
Breath Volume	3.207	15:06
Air Blank	0.000	15:06
Air Blank	0.000	15:08
Subject Sample # 2	0.000	15:09
Breath Volume	1.292	15:09
Air Blank	0.000	15:09
Control Test	0.081	15:10
Air Blank	0.000	15:10
Diagnostics Check	OK	15:10

---

Data Download Date/Time: 11/18/2016 12:30

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: OWEN,SUE,  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: SMITH,STEVEN  
ARRESTING OFFICER AGENCY: Juno Beach PD  
LAST AGENCY INSPECTION: 10/14/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 279870  
DRY GAS EXPIRY DATE: 02/04/2019  
VIOLATION CODE: DUI  
ARREST TIME: 09:11  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 10:28

SUBJECT'S LAST NAME: PINEIRO LETHBRIDGE

DATE/TIME OF TEST: 10/18/2016 11:00:48

---

Diagnostics Check	OK	10:52
Air Blank	0.000	10:52
Control Test	0.081	10:52
Air Blank	0.000	10:53
Subject Sample # 1	0.068	10:55
Breath Volume	1.828	10:55
Air Blank	0.000	10:56
Air Blank	0.000	10:58
Subject Sample # 2	0.065	10:59
Breath Volume	1.593	10:59
Air Blank	0.000	10:59
Control Test	0.080	11:00
Air Blank	0.000	11:00
Diagnostics Check	OK	11:00

---

Data Download Date/Time: 11/18/2016 12:31

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BIGGS,JAMES,G  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: RODRIGUEZ,JOSE  
ARRESTING OFFICER AGENCY: Palm Beach PD  
LAST AGENCY INSPECTION: 10/14/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 279870  
DRY GAS EXPIRY DATE: 02/04/2019  
VIOLATION CODE: DUI  
ARREST TIME: 22:18  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: JIMENEZ DIAZ

DATE/TIME OF TEST: 10/18/2016 23:24:35

---

Diagnostics Check	OK	23:18
Air Blank	0.000	23:18
Control Test	0.081	23:18
Air Blank	0.000	23:19
Subject Sample # 1	0.193	23:19
Breath Volume	2.070	23:19
Air Blank	0.000	23:20
Air Blank	0.000	23:22
Subject Sample # 2	0.191	23:22
Breath Volume	1.214	23:22
Air Blank	0.000	23:23
Control Test	0.079	23:23
Air Blank	0.000	23:24
Diagnostics Check	OK	23:24

---

Data Download Date/Time: 11/18/2016 12:31



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CAIN,JEANETTE,  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: JACKSON,ELI  
ARRESTING OFFICER AGENCY: Riviera Beach PD  
LAST AGENCY INSPECTION: 10/14/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 279870  
DRY GAS EXPIRY DATE: 02/04/2019  
VIOLATION CODE: DUI  
ARREST TIME: 23:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: DORELUS

DATE/TIME OF TEST: 10/19/2016 00:14:10

---

Diagnostics Check		OK	00:07
Air Blank		0.000	00:07
Control Test		0.081	00:07
Air Blank		0.000	00:08
Subject Sample	# 1	0.379	00:09
Breath Volume		1.277	00:09
Air Blank		0.000	00:10
Air Blank		0.000	00:11
Subject Sample	# 2	0.367	00:12
Breath Volume		1.414	00:12
Air Blank		0.000	00:13
Control Test		0.079	00:13
Air Blank		0.000	00:14
Diagnostics Check		OK	00:14

---

Data Download Date/Time: 11/18/2016 12:31

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DORSEY,PAMELA,F  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: LAFORTE,JASON  
ARRESTING OFFICER AGENCY: South Palm Beach  
LAST AGENCY INSPECTION: 10/14/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 279870  
DRY GAS EXPIRY DATE: 02/04/2019  
VIOLATION CODE: DUI  
ARREST TIME: 21:52  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:36

SUBJECT'S LAST NAME: GUSTAFSON

DATE/TIME OF TEST: 10/20/2016 23:18:41

---

Diagnostics Check	OK	23:10
Air Blank	0.000	23:11
Control Test	0.081	23:11
Air Blank	0.000	23:12
Subject Sample # 1	0.191	23:13
Breath Volume	3.550	23:13
Air Blank	0.000	23:14
Air Blank	0.000	23:15
Subject Sample # 2	0.182	23:16
Breath Volume	1.722	23:16
Air Blank	0.000	23:17
Control Test	0.078	23:18
Air Blank	0.000	23:18
Diagnostics Check	OK	23:18

---

Data Download Date/Time: 11/18/2016 12:31

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BIGGS,JAMES,G  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: MERCIER,DANIEL  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 10/14/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 279870  
DRY GAS EXPIRY DATE: 02/04/2019  
VIOLATION CODE: DUI  
ARREST TIME: 01:02  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: OBANDO

DATE/TIME OF TEST: 10/21/2016 02:01:20

---

Diagnostics Check		OK	01:52
Air Blank		0.000	01:52
Control Test		0.081	01:52
Air Blank		0.000	01:53
Subject Sample	# 1	0.205	01:56
Breath Volume		2.207	01:56
Air Blank		0.000	01:57
Air Blank		0.000	01:58
Subject Sample	# 2	0.201	01:59
Breath Volume		1.570	01:59
Air Blank		0.000	02:00
Control Test		0.076	02:00
Air Blank		0.000	02:01
Diagnostics Check		OK	02:01

---

Data Download Date/Time: 11/18/2016 12:31

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DORSEY,PAMELA,F  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: CHIQUITO-  
RODRIGUEZ,LENNDY  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 10/14/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 279870  
DRY GAS EXPIRY DATE: 02/04/2019  
VIOLATION CODE: DUI  
ARREST TIME: 23:24  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: GIL-MUNOZ

DATE/TIME OF TEST: 10/22/2016 00:16:47

---

Diagnostics Check		OK	00:09
Air Blank		0.000	00:09
Control Test		0.081	00:10
Air Blank		0.000	00:10
Subject Sample	# 1	0.164	00:11
Breath Volume		2.742	00:11
Air Blank		0.000	00:12
Air Blank		0.000	00:14
Subject Sample	# 2	0.154	00:14
Breath Volume		2.339	00:14
Air Blank		0.000	00:15
Control Test		0.078	00:16
Air Blank		0.000	00:16
Diagnostics Check		OK	00:16

---

Data Download Date/Time: 11/18/2016 12:31

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: O'NEAL,SHARIL  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: MORAL,L  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 10/14/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 279870  
DRY GAS EXPIRY DATE: 02/04/2019  
VIOLATION CODE: DUI  
ARREST TIME: 01:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: AGUILAR RAMIREZ

DATE/TIME OF TEST: 10/22/2016 03:03:20

---

Diagnostics Check	OK	02:56
Air Blank	0.000	02:57
Control Test	0.081	02:57
Air Blank	0.000	02:58
Subject Sample # 1	0.160	02:58
Breath Volume	2.347	02:58
Air Blank	0.000	02:59
Air Blank	0.000	03:01
Subject Sample # 2	0.161	03:01
Breath Volume	2.312	03:01
Air Blank	0.000	03:02
Control Test	0.079	03:02
Air Blank	0.000	03:03
Diagnostics Check	OK	03:03

---

Data Download Date/Time: 11/18/2016 12:31

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DORSEY,PAMELA,F  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: FICARRA,CHRISTOPHER  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 10/14/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 279870  
DRY GAS EXPIRY DATE: 02/04/2019  
VIOLATION CODE: DUI  
ARREST TIME: 18:45  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 19:47

SUBJECT'S LAST NAME: BURNETT

DATE/TIME OF TEST: 10/22/2016 20:23:47

---

Diagnostics Check	OK	20:15
Air Blank	0.000	20:16
Control Test	0.081	20:16
Air Blank	0.000	20:17
Subject Sample # 1	0.276	20:18
Breath Volume	1.542	20:18
Air Blank	0.000	20:19
Air Blank	0.000	20:20
Subject Sample # 2	0.282	20:21
Breath Volume	2.128	20:21
Air Blank	0.000	20:22
Control Test	0.077	20:23
Air Blank	0.000	20:23
Diagnostics Check	OK	20:23

---

Data Download Date/Time: 11/18/2016 12:31

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DORSEY,PAMELA,F  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: GUARIN,EDGAR  
ARRESTING OFFICER AGENCY: Lantana PD  
LAST AGENCY INSPECTION: 10/14/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 279870  
DRY GAS EXPIRY DATE: 02/04/2019  
VIOLATION CODE: DUI  
ARREST TIME: 21:09  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: BRAZZEAL

DATE/TIME OF TEST: 10/22/2016 22:23:12

---

Diagnostics Check	OK	22:14
Air Blank	0.000	22:15
Control Test	0.080	22:15
Air Blank	0.000	22:16
Subject Sample # 1	0.300	22:18
Breath Volume	1.875	22:18
Air Blank	0.000	22:19
Air Blank	0.000	22:20
Subject Sample # 2	0.295	22:21
Breath Volume	2.492	22:21
Air Blank	0.000	22:22
Control Test	0.076	22:22
Air Blank	0.000	22:23
Diagnostics Check	OK	22:23

---

Data Download Date/Time: 11/18/2016 12:31

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: O'NEAL,SHARIL  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: AVEY,ADAM  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 10/14/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 279870  
DRY GAS EXPIRY DATE: 02/04/2019  
VIOLATION CODE: DUI  
ARREST TIME: 23:41  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: APPLE

DATE/TIME OF TEST: 10/23/2016 00:49:33

---

Diagnostics Check		OK	00:42
Air Blank		0.000	00:42
Control Test		0.080	00:42
Air Blank		0.000	00:43
Subject Sample	# 1	0.162	00:44
Breath Volume		1.191	00:44
Air Blank		0.000	00:45
Air Blank		0.000	00:47
Subject Sample	# 2	0.164	00:47
Breath Volume		1.101	00:47
Air Blank		0.000	00:48
Control Test		0.078	00:48
Air Blank		0.000	00:49
Diagnostics Check		OK	00:49

---

Data Download Date/Time: 11/18/2016 12:31



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DORSEY,PAMELA,F  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: GRUEBEL,KARL  
ARRESTING OFFICER AGENCY: FHP Troop K  
LAST AGENCY INSPECTION: 10/14/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 279870  
DRY GAS EXPIRY DATE: 02/04/2019  
VIOLATION CODE: DUI  
ARREST TIME: 00:33  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: MADDEN

DATE/TIME OF TEST: 10/23/2016 01:58:22

---

Diagnostics Check	OK	01:51
Air Blank	0.000	01:51
Control Test	0.080	01:51
Air Blank	0.000	01:52
Subject Sample # 1	0.171	01:53
Breath Volume	2.722	01:53
Air Blank	0.000	01:54
Air Blank	0.000	01:55
Subject Sample # 2	0.170	01:56
Breath Volume	3.140	01:56
Air Blank	0.000	01:57
Control Test	0.078	01:57
Air Blank	0.000	01:58
Diagnostics Check	OK	01:58

---

Data Download Date/Time: 11/18/2016 12:31

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: OWEN,SUE,  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: SENTMANAT,ADOLFO  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 10/14/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 279870  
DRY GAS EXPIRY DATE: 02/04/2019  
VIOLATION CODE: DUI  
ARREST TIME: 01:39  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: SOMERO

DATE/TIME OF TEST: 10/23/2016 02:41:02

---

Diagnostics Check	OK	02:29
Air Blank	0.000	02:30
Control Test	0.079	02:30
Air Blank	0.000	02:31
Subject Sample # 1	0.209	02:32
Breath Volume	2.058	02:32
Air Blank	0.000	02:33
Air Blank	0.000	02:35
Subject Sample # 2	0.186	02:35
Breath Volume	1.187	02:35
Air Blank	0.000	02:36
Air Blank	0.000	02:38
Subject Sample # 3	0.207	02:39
Breath Volume	2.406	02:39
Air Blank	0.000	02:40
Control Test	0.079	02:40
Air Blank	0.000	02:40
Diagnostics Check	OK	02:41

---

Data Download Date/Time: 11/18/2016 12:31

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: OWEN,SUE,  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: KWAKU,EDWARD  
ARRESTING OFFICER AGENCY: FHP Troop L  
LAST AGENCY INSPECTION: 10/14/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 279870  
DRY GAS EXPIRY DATE: 02/04/2019  
VIOLATION CODE: DUI  
ARREST TIME: 03:16  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: DORVIL

DATE/TIME OF TEST: 10/23/2016 04:31:04

---

Diagnostics Check	OK	04:22
Air Blank	0.000	04:23
Control Test	0.080	04:23
Air Blank	0.000	04:24
Subject Sample # 1	0.153	04:26
Breath Volume	1.117	04:26
Air Blank	0.000	04:27
Air Blank	0.000	04:28
Subject Sample # 2	0.157	04:29
Breath Volume	1.152	04:29
Air Blank	0.000	04:30
Control Test	0.078	04:30
Air Blank	0.000	04:30
Diagnostics Check	OK	04:31

---

Data Download Date/Time: 11/18/2016 12:31

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: OWEN,SUE,  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: MACKKEY,JAMES  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 10/14/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 279870  
DRY GAS EXPIRY DATE: 02/04/2019  
VIOLATION CODE: DUI  
ARREST TIME: 12:37  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 12:55

SUBJECT'S LAST NAME: SANCHEZ-DIAZ

DATE/TIME OF TEST: 10/23/2016 13:43:14

---

Diagnostics Check	OK	13:31	
Air Blank	0.000	13:32	
Control Test	0.081	13:32	
Air Blank	0.000	13:33	
Subject Sample	# 1	13:36	No Sample Provided
Breath Volume	0.000	13:36	
Air Blank	0.000	13:36	
Air Blank	0.000	13:38	
Subject Sample	# 2	13:41	No Sample Provided
Breath Volume	0.000	13:41	
Air Blank	0.000	13:42	
Control Test	0.080	13:42	
Air Blank	0.000	13:43	
Diagnostics Check	OK	13:43	

---

Data Download Date/Time: 11/18/2016 12:31

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: OWEN,SUE,  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: SENTMANAT,ADOLFO  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 10/14/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 279870  
DRY GAS EXPIRY DATE: 02/04/2019  
VIOLATION CODE: DUI  
ARREST TIME: 21:41  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:08

SUBJECT'S LAST NAME: QUINTERO

DATE/TIME OF TEST: 10/23/2016 22:43:17

---

Diagnostics Check		OK	22:35
Air Blank		0.000	22:36
Control Test		0.081	22:36
Air Blank		0.000	22:37
Subject Sample	# 1	0.201	22:38
Breath Volume		2.417	22:38
Air Blank		0.000	22:39
Air Blank		0.000	22:40
Subject Sample	# 2	0.208	22:41
Breath Volume		2.675	22:41
Air Blank		0.000	22:42
Control Test		0.077	22:42
Air Blank		0.000	22:43
Diagnostics Check		OK	22:43

---

Data Download Date/Time: 11/18/2016 12:31

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BIGGS,JAMES,G  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: HOWARD,EMMA  
ARRESTING OFFICER AGENCY: Jupiter PD  
LAST AGENCY INSPECTION: 10/14/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 279870  
DRY GAS EXPIRY DATE: 02/04/2019  
VIOLATION CODE: DUI  
ARREST TIME: 13:21  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 14:12

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 10/26/2016 14:43:28

---

Diagnostics Check	OK	14:36
Air Blank	0.000	14:37
Control Test	0.081	14:37
Air Blank	0.000	14:38
Subject Sample # 1	0.041	14:38
Breath Volume	1.742	14:38
Air Blank	0.000	14:39
Air Blank	0.000	14:41
Subject Sample # 2	0.039	14:41
Breath Volume	1.292	14:41
Air Blank	0.000	14:42
Control Test	0.080	14:42
Air Blank	0.000	14:43
Diagnostics Check	OK	14:43

---

Data Download Date/Time: 11/18/2016 12:31

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BIGGS,JAMES,G  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: PATEL,JESAL  
ARRESTING OFFICER AGENCY: Delray Beach PD  
LAST AGENCY INSPECTION: 10/14/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 279870  
DRY GAS EXPIRY DATE: 02/04/2019  
VIOLATION CODE: DUI  
ARREST TIME: 19:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:30

SUBJECT'S LAST NAME: TRONTI

DATE/TIME OF TEST: 10/27/2016 20:59:55

---

Diagnostics Check	OK	20:53
Air Blank	0.000	20:54
Control Test	0.082	20:54
Air Blank	0.000	20:55
Subject Sample # 1	0.000	20:55
Breath Volume	3.347	20:55
Air Blank	0.000	20:55
Air Blank	0.000	20:57
Subject Sample # 2	0.000	20:58
Breath Volume	3.210	20:58
Air Blank	0.000	20:58
Control Test	0.080	20:59
Air Blank	0.000	20:59
Diagnostics Check	OK	20:59

---

Data Download Date/Time: 11/18/2016 12:31

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: O'NEAL,SHARIL  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: CROY,PAUL  
ARRESTING OFFICER AGENCY: FHP Troop L  
LAST AGENCY INSPECTION: 10/14/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 279870  
DRY GAS EXPIRY DATE: 02/04/2019  
VIOLATION CODE: DUI  
ARREST TIME: 23:21  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: LEWIS

DATE/TIME OF TEST: 10/28/2016 00:37:40

---

Diagnostics Check		OK	00:30
Air Blank		0.000	00:30
Control Test		0.082	00:31
Air Blank		0.000	00:31
Subject Sample	# 1	0.058	00:33
Breath Volume		1.769	00:33
Air Blank		0.000	00:33
Air Blank		0.000	00:35
Subject Sample	# 2	0.063	00:35
Breath Volume		1.550	00:35
Air Blank		0.000	00:36
Control Test		0.079	00:37
Air Blank		0.000	00:37
Diagnostics Check		OK	00:37

---

Data Download Date/Time: 11/18/2016 12:31



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: O'NEAL,SHARIL  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: FICARRA,CHRISTOPHER  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 10/14/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 279870  
DRY GAS EXPIRY DATE: 02/04/2019  
VIOLATION CODE: DUI  
ARREST TIME: 01:17  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:38

SUBJECT'S LAST NAME: LAPOINTE

DATE/TIME OF TEST: 10/29/2016 02:18:05

---

Diagnostics Check	OK	02:09
Air Blank	0.000	02:10
Control Test	0.081	02:10
Air Blank	0.000	02:11
Subject Sample # 1	0.176	02:13
Breath Volume	1.628	02:13
Air Blank	0.000	02:13
Air Blank	0.000	02:15
Subject Sample # 2	0.176	02:16
Breath Volume	1.332	02:16
Air Blank	0.000	02:17
Control Test	0.078	02:17
Air Blank	0.000	02:17
Diagnostics Check	OK	02:18

---

Data Download Date/Time: 11/18/2016 12:31

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: O'NEAL,SHARIL  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: WALTON,THOMAS  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 10/14/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 279870  
DRY GAS EXPIRY DATE: 02/04/2019  
VIOLATION CODE: DUI  
ARREST TIME: 03:27  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:52

SUBJECT'S LAST NAME: HALUSKA

DATE/TIME OF TEST: 10/29/2016 04:24:50

---

Diagnostics Check	OK	04:18
Air Blank	0.000	04:18
Control Test	0.081	04:18
Air Blank	0.000	04:19
Subject Sample # 1	0.220	04:20
Breath Volume	2.562	04:20
Air Blank	0.000	04:20
Air Blank	0.000	04:22
Subject Sample # 2	0.207	04:22
Breath Volume	1.855	04:22
Air Blank	0.000	04:23
Control Test	0.079	04:24
Air Blank	0.000	04:24
Diagnostics Check	OK	04:24

---

Data Download Date/Time: 11/18/2016 12:31

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DORSEY,PAMELA,F  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: STEPHAN,ROBERT  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 10/14/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 279870  
DRY GAS EXPIRY DATE: 02/04/2019  
VIOLATION CODE: DUI  
ARREST TIME: 20:11  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:42

SUBJECT'S LAST NAME: ATIEH

DATE/TIME OF TEST: 10/29/2016 21:17:29

---

Diagnostics Check	OK	21:10
Air Blank	0.000	21:10
Control Test	0.081	21:11
Air Blank	0.000	21:11
Subject Sample # 1	0.000	21:12
Breath Volume	3.792	21:12
Air Blank	0.000	21:13
Air Blank	0.000	21:15
Subject Sample # 2	0.000	21:16
Breath Volume	4.082	21:16
Air Blank	0.000	21:16
Control Test	0.080	21:16
Air Blank	0.000	21:17
Diagnostics Check	OK	21:17

---

Data Download Date/Time: 11/18/2016 12:31

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DORSEY,PAMELA,F  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: CROY,PAUL  
ARRESTING OFFICER AGENCY: FHP Troop L  
LAST AGENCY INSPECTION: 10/14/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 279870  
DRY GAS EXPIRY DATE: 02/04/2019  
VIOLATION CODE: DUI  
ARREST TIME: 21:24  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:21

SUBJECT'S LAST NAME: MORGAN

DATE/TIME OF TEST: 10/29/2016 23:07:12

---

Diagnostics Check	OK	22:59
Air Blank	0.000	23:00
Control Test	0.081	23:00
Air Blank	0.000	23:01
Subject Sample # 1	0.130	23:02
Breath Volume	1.785	23:02
Air Blank	0.000	23:03
Air Blank	0.000	23:04
Subject Sample # 2	0.130	23:05
Breath Volume	1.980	23:05
Air Blank	0.000	23:06
Control Test	0.079	23:06
Air Blank	0.000	23:07
Diagnostics Check	OK	23:07

---

Data Download Date/Time: 11/18/2016 12:31

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DORSEY,PAMELA,F  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: GRUEBEL,KARL  
ARRESTING OFFICER AGENCY: FHP Troop K  
LAST AGENCY INSPECTION: 10/14/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 279870  
DRY GAS EXPIRY DATE: 02/04/2019  
VIOLATION CODE: DUI  
ARREST TIME: 12:52  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: HERNANDEZ-ESCOBAR

DATE/TIME OF TEST: 10/30/2016 02:12:31

---

Diagnostics Check	OK	02:04
Air Blank	0.000	02:05
Control Test	0.081	02:05
Air Blank	0.000	02:06
Subject Sample # 1	0.163	02:07
Breath Volume	1.691	02:07
Air Blank	0.000	02:08
Air Blank	0.000	02:09
Subject Sample # 2	0.159	02:10
Breath Volume	1.265	02:10
Air Blank	0.000	02:11
Control Test	0.078	02:11
Air Blank	0.000	02:12
Diagnostics Check	OK	02:12

---

Data Download Date/Time: 11/18/2016 12:31

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DORSEY,PAMELA,F  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: ZESUT,BRANDEN  
ARRESTING OFFICER AGENCY: Jupiter PD  
LAST AGENCY INSPECTION: 10/14/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 279870  
DRY GAS EXPIRY DATE: 02/04/2019  
VIOLATION CODE: DUI  
ARREST TIME: 21:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:57

SUBJECT'S LAST NAME: CUUCORE

DATE/TIME OF TEST: 10/30/2016 22:36:14

---

Diagnostics Check	OK	22:27	
Air Blank	0.000	22:27	
Control Test	0.081	22:27	
Air Blank	0.000	22:28	
Subject Sample	# 1	22:31	No Sample Provided
Breath Volume	0.000	22:31	
Air Blank	0.000	22:32	
Air Blank	0.000	22:34	
Subject Sample	# 2	22:34	Subject Test Refused
Breath Volume	0.000	22:34	
Air Blank	0.000	22:35	
Control Test	0.080	22:35	
Air Blank	0.000	22:36	
Diagnostics Check	OK	22:36	

---

Data Download Date/Time: 11/18/2016 12:31

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DORSEY,PAMELA,F  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: GEE,JOHN  
ARRESTING OFFICER AGENCY: Palm Springs PD  
LAST AGENCY INSPECTION: 09/09/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 279870  
DRY GAS EXPIRY DATE: 02/04/2019  
VIOLATION CODE: DUI  
ARREST TIME: 22:18  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 10/01/2016 23:33:09

---

Diagnostics Check		OK	23:21	
Air Blank		0.000	23:21	
Control Test		0.081	23:21	
Air Blank		0.000	23:22	
Subject Sample	# 1	0.200	23:25	Volume Not Met
Breath Volume		0.070	23:25	
Air Blank		0.000	23:26	
Air Blank		0.000	23:28	
Subject Sample	# 2	0.177	23:31	Volume Not Met
Breath Volume		0.421	23:31	
Air Blank		0.000	23:32	
Control Test		0.075	23:32	
Air Blank		0.000	23:33	
Diagnostics Check		OK	23:33	

---

Data Download Date/Time: 10/14/2016 12:23

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: OWEN,SUE,  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: PUJOL,JESUS  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 09/09/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 279870  
DRY GAS EXPIRY DATE: 02/04/2019  
VIOLATION CODE: DUI  
ARREST TIME: 23:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:28

SUBJECT'S LAST NAME: ESCALANTE

DATE/TIME OF TEST: 10/02/2016 00:59:15

---

Diagnostics Check		OK	00:51
Air Blank		0.000	00:52
Control Test		0.078	00:52
Air Blank		0.000	00:52
Subject Sample	# 1	0.183	00:54
Breath Volume		2.000	00:54
Air Blank		0.000	00:55
Air Blank		0.000	00:56
Subject Sample	# 2	0.182	00:57
Breath Volume		1.250	00:57
Air Blank		0.000	00:58
Control Test		0.078	00:58
Air Blank		0.000	00:59
Diagnostics Check		OK	00:59

---

Data Download Date/Time: 10/14/2016 12:23



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: OWEN,SUE,  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: SCHNEIDER,JOSHUA  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 09/09/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 279870  
DRY GAS EXPIRY DATE: 02/04/2019  
VIOLATION CODE: DUI  
ARREST TIME: 07:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 08:14

SUBJECT'S LAST NAME: STEVENS

DATE/TIME OF TEST: 10/03/2016 08:47:21

---

Diagnostics Check	OK	08:41
Air Blank	0.000	08:41
Control Test	0.082	08:41
Air Blank	0.000	08:42
Subject Sample # 1	0.000	08:43
Breath Volume	1.222	08:43
Air Blank	0.000	08:43
Air Blank	0.000	08:45
Subject Sample # 2	0.000	08:45
Breath Volume	1.445	08:45
Air Blank	0.000	08:46
Control Test	0.081	08:46
Air Blank	0.000	08:47
Diagnostics Check	OK	08:47

---

Data Download Date/Time: 10/14/2016 12:23

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: OWEN,SUE,  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: HUYETT,MARK  
ARRESTING OFFICER AGENCY: Delray Beach PD  
LAST AGENCY INSPECTION: 09/09/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 279870  
DRY GAS EXPIRY DATE: 02/04/2019  
VIOLATION CODE: DUI  
ARREST TIME: 14:38  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 15:07

SUBJECT'S LAST NAME: DOUBLER

DATE/TIME OF TEST: 10/05/2016 15:38:24

---

Diagnostics Check	OK	15:31
Air Blank	0.000	15:32
Control Test	0.081	15:32
Air Blank	0.000	15:33
Subject Sample # 1	0.260	15:33
Breath Volume	2.128	15:33
Air Blank	0.000	15:34
Air Blank	0.000	15:36
Subject Sample # 2	0.267	15:36
Breath Volume	2.492	15:36
Air Blank	0.000	15:37
Control Test	0.080	15:37
Air Blank	0.000	15:38
Diagnostics Check	OK	15:38

---

Data Download Date/Time: 10/14/2016 12:23

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: OWEN,SUE,  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: FLEMING,JOHN  
ARRESTING OFFICER AGENCY: Juno Beach PD  
LAST AGENCY INSPECTION: 09/09/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 279870  
DRY GAS EXPIRY DATE: 02/04/2019  
VIOLATION CODE: DUI  
ARREST TIME: 15:12  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 16:08

SUBJECT'S LAST NAME: LAMBROU

DATE/TIME OF TEST: 10/05/2016 16:36:55

---

Diagnostics Check	OK	16:30	
Air Blank	0.000	16:31	
Control Test	0.081	16:31	
Air Blank	0.000	16:32	
Subject Sample	# 1	16:35	Subject Test Refused
Breath Volume	0.394	16:35	
Air Blank	0.000	16:35	
Control Test	0.081	16:36	
Air Blank	0.000	16:36	
Diagnostics Check	OK	16:36	

---

Data Download Date/Time: 10/14/2016 12:23

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DORSEY,PAMELA,F  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: CLAUDIO,JOSE  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 09/09/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 279870  
DRY GAS EXPIRY DATE: 02/04/2019  
VIOLATION CODE: DUI  
ARREST TIME: 12:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 14:19

SUBJECT'S LAST NAME: GARCIA-FRANCO

DATE/TIME OF TEST: 10/07/2016 14:53:46

---

Diagnostics Check	OK	14:46
Air Blank	0.000	14:46
Control Test	0.082	14:47
Air Blank	0.000	14:47
Subject Sample # 1	0.142	14:48
Breath Volume	1.300	14:48
Air Blank	0.000	14:49
Air Blank	0.000	14:51
Subject Sample # 2	0.147	14:52
Breath Volume	1.839	14:52
Air Blank	0.000	14:52
Control Test	0.079	14:53
Air Blank	0.000	14:53
Diagnostics Check	OK	14:53

---

Data Download Date/Time: 10/14/2016 12:23

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: OWEN,SUE,  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: SCHNEIDER,JOSHUA  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 09/09/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 279870  
DRY GAS EXPIRY DATE: 02/04/2019  
VIOLATION CODE: DUI  
ARREST TIME: 09:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 10:05

SUBJECT'S LAST NAME: PADILLA

DATE/TIME OF TEST: 10/08/2016 11:08:30

---

Diagnostics Check	OK	11:01
Air Blank	0.000	11:01
Control Test	0.081	11:02
Air Blank	0.000	11:02
Subject Sample # 1	0.229	11:03
Breath Volume	1.921	11:03
Air Blank	0.000	11:04
Air Blank	0.000	11:06
Subject Sample # 2	0.232	11:06
Breath Volume	1.910	11:06
Air Blank	0.000	11:07
Control Test	0.080	11:07
Air Blank	0.000	11:08
Diagnostics Check	OK	11:08

---

Data Download Date/Time: 10/14/2016 12:23

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DORSEY,PAMELA,F  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: MANNING,WILLIAM  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 09/09/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 279870  
DRY GAS EXPIRY DATE: 02/04/2019  
VIOLATION CODE: DUI  
ARREST TIME: 16:22  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:53

SUBJECT'S LAST NAME: DELAY

DATE/TIME OF TEST: 10/08/2016 19:14:25

---

Diagnostics Check	OK	19:06
Air Blank	0.000	19:06
Control Test	0.081	19:07
Air Blank	0.000	19:07
Subject Sample # 1	0.119	19:09
Breath Volume	1.253	19:09
Air Blank	0.000	19:10
Air Blank	0.000	19:11
Subject Sample # 2	0.134	19:12
Breath Volume	2.156	19:12
Air Blank	0.000	19:13
Control Test	0.081	19:13
Air Blank	0.000	19:14
Diagnostics Check	OK	19:14

---

Data Download Date/Time: 10/14/2016 12:23

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DORSEY,PAMELA,F  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: MORA,AARON  
ARRESTING OFFICER AGENCY: FHP Troop K  
LAST AGENCY INSPECTION: 09/09/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 279870  
DRY GAS EXPIRY DATE: 02/04/2019  
VIOLATION CODE: DUI  
ARREST TIME: 19:29  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:20

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 10/08/2016 21:01:50

---

Diagnostics Check	OK	20:54
Air Blank	0.000	20:55
Control Test	0.081	20:55
Air Blank	0.000	20:56
Subject Sample # 1	0.170	20:56
Breath Volume	1.914	20:56
Air Blank	0.000	20:57
Air Blank	0.000	20:59
Subject Sample # 2	0.172	21:00
Breath Volume	2.589	21:00
Air Blank	0.000	21:00
Control Test	0.081	21:01
Air Blank	0.000	21:01
Diagnostics Check	OK	21:01

---

Data Download Date/Time: 10/14/2016 12:23

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DORSEY,PAMELA,F  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: HANTON,MELINDA  
ARRESTING OFFICER AGENCY: Palm Beach Gardens  
LAST AGENCY INSPECTION: 09/09/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 279870  
DRY GAS EXPIRY DATE: 02/04/2019  
VIOLATION CODE: DUI  
ARREST TIME: 20:14  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 20:47

SUBJECT'S LAST NAME: GENIS

DATE/TIME OF TEST: 10/08/2016 21:36:58

---

Diagnostics Check	OK	21:28
Air Blank	0.000	21:28
Control Test	0.082	21:28
Air Blank	0.000	21:29
Subject Sample # 1	0.190	21:32
Breath Volume	1.574	21:32
Air Blank	0.000	21:32
Air Blank	0.000	21:34
Subject Sample # 2	0.202	21:35
Breath Volume	1.839	21:35
Air Blank	0.000	21:36
Control Test	0.081	21:36
Air Blank	0.000	21:36
Diagnostics Check	OK	21:36

---

Data Download Date/Time: 10/14/2016 12:23



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DORSEY,PAMELA,F  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: DASILVA,MARTIN  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 09/09/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 279870  
DRY GAS EXPIRY DATE: 02/04/2019  
VIOLATION CODE: DUI  
ARREST TIME: 19:34  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:10

SUBJECT'S LAST NAME: FILIPONE

DATE/TIME OF TEST: 10/09/2016 20:41:26

---

Diagnostics Check	OK	20:34
Air Blank	0.000	20:35
Control Test	0.082	20:35
Air Blank	0.000	20:36
Subject Sample # 1	0.000	20:36
Breath Volume	2.738	20:36
Air Blank	0.000	20:37
Air Blank	0.000	20:39
Subject Sample # 2	0.000	20:40
Breath Volume	1.968	20:40
Air Blank	0.000	20:40
Control Test	0.081	20:40
Air Blank	0.000	20:41
Diagnostics Check	OK	20:41

---

Data Download Date/Time: 10/14/2016 12:23

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: OWEN,SUE,  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: PUJOL,JESUS  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 09/09/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 279870  
DRY GAS EXPIRY DATE: 02/04/2019  
VIOLATION CODE: DUI  
ARREST TIME: 00:30  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:03

SUBJECT'S LAST NAME: OBER

DATE/TIME OF TEST: 10/10/2016 01:33:02

---

Diagnostics Check	OK	01:26
Air Blank	0.000	01:26
Control Test	0.081	01:26
Air Blank	0.000	01:27
Subject Sample # 1	0.211	01:28
Breath Volume	2.628	01:28
Air Blank	0.000	01:28
Air Blank	0.000	01:30
Subject Sample # 2	0.225	01:31
Breath Volume	1.734	01:31
Air Blank	0.000	01:32
Control Test	0.080	01:32
Air Blank	0.000	01:32
Diagnostics Check	OK	01:33

---

Data Download Date/Time: 10/14/2016 12:23

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BIGGS,JAMES,G  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: LEVEY,SCOTT  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 09/09/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 279870  
DRY GAS EXPIRY DATE: 02/04/2019  
VIOLATION CODE: DUI  
ARREST TIME: 01:49  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: KAMEN

DATE/TIME OF TEST: 10/12/2016 03:04:18

---

Diagnostics Check	OK	02:58
Air Blank	0.000	02:58
Control Test	0.082	02:58
Air Blank	0.000	02:59
Subject Sample # 1	0.071	02:59
Breath Volume	1.179	02:59
Air Blank	0.000	03:00
Air Blank	0.000	03:02
Subject Sample # 2	0.074	03:02
Breath Volume	2.355	03:02
Air Blank	0.000	03:03
Control Test	0.082	03:03
Air Blank	0.000	03:04
Diagnostics Check	OK	03:04

---

Data Download Date/Time: 10/14/2016 12:23

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BIGGS,JAMES,G  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: FANDREY,CHRISTOPHER  
ARRESTING OFFICER AGENCY: Jupiter PD  
LAST AGENCY INSPECTION: 09/09/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 279870  
DRY GAS EXPIRY DATE: 02/04/2019  
VIOLATION CODE: DUI  
ARREST TIME: 02:42  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:19

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 10/14/2016 04:49:18

---

Diagnostics Check	OK	04:43
Air Blank	0.000	04:43
Control Test	0.081	04:43
Air Blank	0.000	04:44
Subject Sample # 1	0.142	04:44
Breath Volume	1.289	04:44
Air Blank	0.000	04:45
Air Blank	0.000	04:47
Subject Sample # 2	0.141	04:47
Breath Volume	1.164	04:47
Air Blank	0.000	04:48
Control Test	0.081	04:48
Air Blank	0.000	04:49
Diagnostics Check	OK	04:49

---

Data Download Date/Time: 10/14/2016 12:23

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: O'NEAL,SHARIL  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: HANTON,MELINDA  
ARRESTING OFFICER AGENCY: Palm Beach Gardens  
LAST AGENCY INSPECTION: 10/14/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 279870  
DRY GAS EXPIRY DATE: 02/04/2019  
VIOLATION CODE: DUI  
ARREST TIME: 00:03  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: HARVIN

DATE/TIME OF TEST: 10/16/2016 01:08:48

---

Diagnostics Check	OK	01:02
Air Blank	0.000	01:02
Control Test	0.081	01:03
Air Blank	0.000	01:03
Subject Sample # 1	0.136	01:04
Breath Volume	2.468	01:04
Air Blank	0.000	01:04
Air Blank	0.000	01:06
Subject Sample # 2	0.128	01:07
Breath Volume	2.289	01:07
Air Blank	0.000	01:07
Control Test	0.081	01:08
Air Blank	0.000	01:08
Diagnostics Check	OK	01:08

---

Data Download Date/Time: 11/18/2016 16:11

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PARENT,GARY,J  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: WHITE,EDWARD  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 10/14/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 279870  
DRY GAS EXPIRY DATE: 02/04/2019  
VIOLATION CODE: DUI  
ARREST TIME: 19:58  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 20:23

SUBJECT'S LAST NAME: PORTALATIN

DATE/TIME OF TEST: 10/16/2016 21:34:11

---

Diagnostics Check	OK	21:26
Air Blank	0.000	21:27
Control Test	0.082	21:27
Air Blank	0.000	21:28
Subject Sample # 1	0.180	21:29
Breath Volume	1.972	21:29
Air Blank	0.000	21:30
Air Blank	0.000	21:32
Subject Sample # 2	0.169	21:32
Breath Volume	1.511	21:32
Air Blank	0.000	21:33
Control Test	0.080	21:33
Air Blank	0.000	21:34
Diagnostics Check	OK	21:34

---

Data Download Date/Time: 11/18/2016 16:11

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CAIN,JEANETTE,  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: GUARIN,EDGAR  
ARRESTING OFFICER AGENCY: Lantana PD  
LAST AGENCY INSPECTION: 10/14/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 279870  
DRY GAS EXPIRY DATE: 02/04/2019  
VIOLATION CODE: DUI  
ARREST TIME: 00:03  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: HOLCOMB

DATE/TIME OF TEST: 10/18/2016 01:01:50

---

Diagnostics Check	OK	00:58	
Air Blank	0.000	00:59	
Control Test	0.082	00:59	
Air Blank	0.000	01:00	
Subject Sample	# 1	01:00	Subject Test Refused
Breath Volume	0.000	01:00	
Air Blank	0.000	01:00	
Control Test	0.081	01:01	
Air Blank	0.000	01:01	
Diagnostics Check	OK	01:01	

---

Data Download Date/Time: 11/18/2016 16:11

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PARENT,GARY,J  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: BORROWS,ANDREW  
ARRESTING OFFICER AGENCY: Jupiter PD  
LAST AGENCY INSPECTION: 10/14/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 279870  
DRY GAS EXPIRY DATE: 02/04/2019  
VIOLATION CODE: DUI  
ARREST TIME: 03:31  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 10/18/2016 04:53:39

---

Diagnostics Check	OK	04:46
Air Blank	0.000	04:46
Control Test	0.082	04:47
Air Blank	0.000	04:47
Subject Sample # 1	0.157	04:48
Breath Volume	2.996	04:48
Air Blank	0.000	04:49
Air Blank	0.000	04:51
Subject Sample # 2	0.155	04:51
Breath Volume	2.636	04:51
Air Blank	0.000	04:52
Control Test	0.081	04:53
Air Blank	0.000	04:53
Diagnostics Check	OK	04:53

---

Data Download Date/Time: 11/18/2016 16:11



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CAIN,JEANETTE,  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: CHIQUITO-  
RODRIGUEZ,LENNDY  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 10/14/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 279870  
DRY GAS EXPIRY DATE: 02/04/2019  
VIOLATION CODE: DUI  
ARREST TIME: 18:51  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 19:15

SUBJECT'S LAST NAME: CHOKANIS

DATE/TIME OF TEST: 10/18/2016 19:49:27

---

Diagnostics Check		OK	19:41
Air Blank		0.000	19:42
Control Test		0.082	19:42
Air Blank		0.000	19:42
Subject Sample	# 1	0.154	19:44
Breath Volume		1.296	19:44
Air Blank		0.000	19:45
Air Blank		0.000	19:47
Subject Sample	# 2	0.156	19:47
Breath Volume		1.503	19:47
Air Blank		0.000	19:48
Control Test		0.080	19:48
Air Blank		0.000	19:49
Diagnostics Check		OK	19:49

---

Data Download Date/Time: 11/18/2016 16:11

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PARENT,GARY,J  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: MOREA,DEAN  
ARRESTING OFFICER AGENCY: Palm Beach PD  
LAST AGENCY INSPECTION: 10/14/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 279870  
DRY GAS EXPIRY DATE: 02/04/2019  
VIOLATION CODE: DUI  
ARREST TIME: 23:09  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: GUSTAFSON

DATE/TIME OF TEST: 10/19/2016 00:03:52

---

Diagnostics Check	OK	23:56
Air Blank	0.000	23:57
Control Test	0.080	23:57
Air Blank	0.000	23:58
Subject Sample # 2	0.142	23:59
Breath Volume	1.671	23:59
Air Blank	0.000	23:59
Air Blank	0.000	00:01
Subject Sample # 1	0.141	00:02
Breath Volume	1.488	00:02
Air Blank	0.000	00:02
Control Test	0.081	00:03
Air Blank	0.000	00:03
Diagnostics Check	OK	00:03

---

Data Download Date/Time: 11/18/2016 16:11

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: O'NEAL,SHARIL  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: SAYIH,ROBERT  
ARRESTING OFFICER AGENCY: FHP Troop L  
LAST AGENCY INSPECTION: 10/14/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 279870  
DRY GAS EXPIRY DATE: 02/04/2019  
VIOLATION CODE: DUI  
ARREST TIME: 00:46  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:04

SUBJECT'S LAST NAME: TYLER

DATE/TIME OF TEST: 10/20/2016 01:45:27

---

Diagnostics Check	OK	01:36
Air Blank	0.000	01:36
Control Test	0.082	01:36
Air Blank	0.000	01:37
Subject Sample # 1	0.199	01:40
Breath Volume	1.398	01:40
Air Blank	0.000	01:41
Air Blank	0.000	01:42
Subject Sample # 2	0.196	01:43
Breath Volume	1.484	01:43
Air Blank	0.000	01:44
Control Test	0.079	01:44
Air Blank	0.000	01:45
Diagnostics Check	OK	01:45

---

Data Download Date/Time: 11/18/2016 16:11

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: O'NEAL,SHARIL  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: BORROWS,ANDREW  
ARRESTING OFFICER AGENCY: Jupiter PD  
LAST AGENCY INSPECTION: 10/14/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 279870  
DRY GAS EXPIRY DATE: 02/04/2019  
VIOLATION CODE: DUI  
ARREST TIME: 19:50  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 20:28

SUBJECT'S LAST NAME: CARGILL

DATE/TIME OF TEST: 10/20/2016 21:03:39

---

Diagnostics Check	OK	20:53
Air Blank	0.000	20:53
Control Test	0.081	20:54
Air Blank	0.000	20:54
Subject Sample # 1	0.071	20:57
Breath Volume	1.414	20:57
Air Blank	0.000	20:58
Air Blank	0.000	21:00
Subject Sample # 2	0.071	21:02
Breath Volume	1.132	21:02
Air Blank	0.000	21:02
Control Test	0.080	21:03
Air Blank	0.000	21:03
Diagnostics Check	OK	21:03

---

Data Download Date/Time: 11/18/2016 16:11

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: O'NEAL,SHARIL  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: FREY,JACOB  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 10/14/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 279870  
DRY GAS EXPIRY DATE: 02/04/2019  
VIOLATION CODE: DUI  
ARREST TIME: 00:49  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:47

SUBJECT'S LAST NAME: ZARA

DATE/TIME OF TEST: 10/21/2016 02:26:18

---

Diagnostics Check	OK	02:14	
Air Blank	0.000	02:14	
Control Test	0.081	02:15	
Air Blank	0.000	02:15	
Subject Sample	# 1 0.191	02:18	Volume Not Met
Breath Volume	0.046	02:18	
Air Blank	0.000	02:19	
Air Blank	0.000	02:21	
Subject Sample	# 2 0.187	02:24	Volume Not Met
Breath Volume	0.097	02:24	
Air Blank	0.000	02:25	
Control Test	0.079	02:25	
Air Blank	0.000	02:26	
Diagnostics Check	OK	02:26	

---

Data Download Date/Time: 11/18/2016 16:11

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: O'NEAL,SHARIL  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: FREY,JACOB  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 10/14/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 279870  
DRY GAS EXPIRY DATE: 02/04/2019  
VIOLATION CODE: DUI  
ARREST TIME: 00:49  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:47

SUBJECT'S LAST NAME: ZARA

DATE/TIME OF TEST: 10/21/2016 02:36:10

---

Diagnostics Check		OK	02:27	
Air Blank		0.000	02:28	
Control Test		0.079	02:28	
Air Blank		0.000	02:29	
Subject Sample	# 1	0.187	02:32	Volume Not Met
Breath Volume		0.070	02:32	
Air Blank		0.000	02:32	
Air Blank		0.000	02:34	
Subject Sample	# 2		02:34	Subject Test Refused
Breath Volume		0.000	02:34	
Air Blank		0.000	02:35	
Control Test		0.077	02:35	
Air Blank		0.000	02:36	
Diagnostics Check		OK	02:36	

---

Data Download Date/Time: 11/18/2016 16:11

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: OWEN,SUE,  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: SNOW,JOHN  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 10/14/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 279870  
DRY GAS EXPIRY DATE: 02/04/2019  
VIOLATION CODE: DUI  
ARREST TIME: 23:44  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:26

SUBJECT'S LAST NAME: TUCKER

DATE/TIME OF TEST: 10/22/2016 01:07:49

---

Diagnostics Check	OK	00:55	
Air Blank	0.000	00:55	
Control Test	0.081	00:55	
Air Blank	0.000	00:56	
Subject Sample	# 1 0.130	00:57	
Breath Volume	1.144	00:57	
Air Blank	0.000	00:58	
Air Blank	0.000	00:59	
Subject Sample	# 2 0.122	01:03	Volume Not Met
Breath Volume	0.183	01:03	
Air Blank	0.000	01:03	
Air Blank	0.000	01:05	
Subject Sample	# 3 0.127	01:06	
Breath Volume	1.125	01:06	
Air Blank	0.000	01:06	
Control Test	0.079	01:07	
Air Blank	0.000	01:07	
Diagnostics Check	OK	01:07	

---

Data Download Date/Time: 11/18/2016 16:11

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: OWEN,SUE,  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: SAYIH,ROBERT  
ARRESTING OFFICER AGENCY: FHP Troop L  
LAST AGENCY INSPECTION: 10/14/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 279870  
DRY GAS EXPIRY DATE: 02/04/2019  
VIOLATION CODE: DUI  
ARREST TIME: 02:29  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:11

SUBJECT'S LAST NAME: HERRERA

DATE/TIME OF TEST: 10/22/2016 03:46:25

---

Diagnostics Check		OK	03:38
Air Blank		0.000	03:39
Control Test		0.081	03:39
Air Blank		0.000	03:40
Subject Sample	# 1	0.269	03:41
Breath Volume		2.890	03:41
Air Blank		0.000	03:42
Air Blank		0.000	03:44
Subject Sample	# 2	0.254	03:44
Breath Volume		2.078	03:44
Air Blank		0.000	03:45
Control Test		0.079	03:45
Air Blank		0.000	03:46
Diagnostics Check		OK	03:46

---

Data Download Date/Time: 11/18/2016 16:11



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: OWEN,SUE,  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: CROY,PAUL  
ARRESTING OFFICER AGENCY: FHP Troop L  
LAST AGENCY INSPECTION: 10/14/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 279870  
DRY GAS EXPIRY DATE: 02/04/2019  
VIOLATION CODE: DUI  
ARREST TIME: 03:39  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:43

SUBJECT'S LAST NAME: WALTERS

DATE/TIME OF TEST: 10/22/2016 05:14:44

---

Diagnostics Check	OK	05:08
Air Blank	0.000	05:08
Control Test	0.081	05:08
Air Blank	0.000	05:09
Subject Sample # 1	0.169	05:10
Breath Volume	1.640	05:10
Air Blank	0.000	05:10
Air Blank	0.000	05:12
Subject Sample # 2	0.161	05:13
Breath Volume	1.218	05:13
Air Blank	0.000	05:13
Control Test	0.080	05:14
Air Blank	0.000	05:14
Diagnostics Check	OK	05:14

---

Data Download Date/Time: 11/18/2016 16:11

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: OWEN,SUE,  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: SCHNEIDER,JOSHUA  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 10/14/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 279870  
DRY GAS EXPIRY DATE: 02/04/2019  
VIOLATION CODE: DUI  
ARREST TIME: 06:51  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 07:16

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 10/22/2016 07:49:30

---

Diagnostics Check		OK	07:42
Air Blank		0.000	07:42
Control Test		0.081	07:43
Air Blank		0.000	07:43
Subject Sample	# 1	0.155	07:44
Breath Volume		2.636	07:44
Air Blank		0.000	07:45
Air Blank		0.000	07:47
Subject Sample	# 2	0.164	07:47
Breath Volume		2.804	07:47
Air Blank		0.000	07:48
Control Test		0.081	07:48
Air Blank		0.000	07:49
Diagnostics Check		OK	07:49

---

Data Download Date/Time: 11/18/2016 16:11

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: OWEN,SUE,  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: TOBY,BRADLEY  
ARRESTING OFFICER AGENCY: FFWCC (West Palm)  
LAST AGENCY INSPECTION: 10/14/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 279870  
DRY GAS EXPIRY DATE: 02/04/2019  
VIOLATION CODE: DUI  
ARREST TIME: 17:53  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:24

SUBJECT'S LAST NAME: CAMELLO RUBIO

DATE/TIME OF TEST: 10/23/2016 00:59:09

---

Diagnostics Check	OK	00:52
Air Blank	0.000	00:53
Control Test	0.082	00:53
Air Blank	0.000	00:54
Subject Sample # 1	0.085	00:54
Breath Volume	1.273	00:54
Air Blank	0.000	00:55
Air Blank	0.000	00:57
Subject Sample # 2	0.085	00:57
Breath Volume	1.410	00:57
Air Blank	0.000	00:58
Control Test	0.081	00:58
Air Blank	0.000	00:59
Diagnostics Check	OK	00:59

---

Data Download Date/Time: 11/18/2016 16:11

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: O'NEAL,SHARIL  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: SAYIH,ROBERT  
ARRESTING OFFICER AGENCY: FHP Troop L  
LAST AGENCY INSPECTION: 10/14/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 279870  
DRY GAS EXPIRY DATE: 02/04/2019  
VIOLATION CODE: DUI  
ARREST TIME: 01:04  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:38

SUBJECT'S LAST NAME: ALVAREZ GALICIA

DATE/TIME OF TEST: 10/23/2016 02:20:31

---

Diagnostics Check		OK	02:08	
Air Blank		0.000	02:09	
Control Test		0.080	02:09	
Air Blank		0.000	02:09	
Subject Sample	# 1	0.170	02:13	Volume Not Met
Breath Volume		0.226	02:13	
Air Blank		0.000	02:13	
Air Blank		0.000	02:15	
Subject Sample	# 2	0.190	02:18	Volume Not Met
Breath Volume		0.882	02:18	
Air Blank		0.000	02:19	
Control Test		0.079	02:19	
Air Blank		0.000	02:20	
Diagnostics Check		OK	02:20	

---

Data Download Date/Time: 11/18/2016 16:11

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: O'NEAL,SHARIL  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: SAYIH,ROBERT  
ARRESTING OFFICER AGENCY: FHP Troop L  
LAST AGENCY INSPECTION: 10/14/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 279870  
DRY GAS EXPIRY DATE: 02/04/2019  
VIOLATION CODE: DUI  
ARREST TIME: 01:04  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:38

SUBJECT'S LAST NAME: ALVAREZ GALICIA

DATE/TIME OF TEST: 10/23/2016 02:27:55

---

Diagnostics Check	OK	02:21	
Air Blank	0.000	02:22	
Control Test	0.079	02:22	
Air Blank	0.000	02:23	
Subject Sample	# 1	02:26	Subject Test Refused
Breath Volume	0.390	02:26	
Air Blank	0.000	02:26	
Control Test	0.079	02:27	
Air Blank	0.000	02:27	
Diagnostics Check	OK	02:27	

---

Data Download Date/Time: 11/18/2016 16:11

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: O'NEAL,SHARIL  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: SERRAES,BRANDON  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 10/14/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 279870  
DRY GAS EXPIRY DATE: 02/04/2019  
VIOLATION CODE: DUI  
ARREST TIME: 01:03  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:08

SUBJECT'S LAST NAME: CREAMER

DATE/TIME OF TEST: 10/23/2016 02:54:10

---

Diagnostics Check	OK	02:47
Air Blank	0.000	02:48
Control Test	0.079	02:48
Air Blank	0.000	02:49
Subject Sample # 1	0.191	02:49
Breath Volume	1.656	02:49
Air Blank	0.000	02:50
Air Blank	0.000	02:52
Subject Sample # 2	0.186	02:52
Breath Volume	1.320	02:52
Air Blank	0.000	02:53
Control Test	0.079	02:53
Air Blank	0.000	02:54
Diagnostics Check	OK	02:54

---

Data Download Date/Time: 11/18/2016 16:11

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: O'NEAL,SHARIL  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: WEBB,ALAN  
ARRESTING OFFICER AGENCY: FFWCC (West Palm)  
LAST AGENCY INSPECTION: 10/14/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 279870  
DRY GAS EXPIRY DATE: 02/04/2019  
VIOLATION CODE: DUI  
ARREST TIME: 02:13  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: DOOLITTLE BRUNO

DATE/TIME OF TEST: 10/23/2016 04:11:26

---

Diagnostics Check	OK	04:04
Air Blank	0.000	04:05
Control Test	0.081	04:05
Air Blank	0.000	04:06
Subject Sample # 1	0.157	04:06
Breath Volume	2.152	04:06
Air Blank	0.000	04:07
Air Blank	0.000	04:09
Subject Sample # 2	0.153	04:09
Breath Volume	1.339	04:09
Air Blank	0.000	04:10
Control Test	0.081	04:10
Air Blank	0.000	04:11
Diagnostics Check	OK	04:11

---

Data Download Date/Time: 11/18/2016 16:11

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: O'NEAL,SHARIL  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: ARANGO,JOSE  
ARRESTING OFFICER AGENCY: Jupiter PD  
LAST AGENCY INSPECTION: 10/14/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 279870  
DRY GAS EXPIRY DATE: 02/04/2019  
VIOLATION CODE: DUI  
ARREST TIME: 07:28  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 07:54

SUBJECT'S LAST NAME: AILON MENDOZA

DATE/TIME OF TEST: 10/23/2016 08:28:36

---

Diagnostics Check		OK	08:21
Air Blank		0.000	08:21
Control Test		0.082	08:22
Air Blank		0.000	08:22
Subject Sample	# 1	0.200	08:23
Breath Volume		2.406	08:23
Air Blank		0.000	08:24
Air Blank		0.000	08:25
Subject Sample	# 2	0.207	08:26
Breath Volume		1.585	08:26
Air Blank		0.000	08:27
Control Test		0.081	08:27
Air Blank		0.000	08:28
Diagnostics Check		OK	08:28

---

Data Download Date/Time: 11/18/2016 16:11



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CAIN,JEANETTE,  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: SENTMANAT,ADOLFO  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 10/14/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 279870  
DRY GAS EXPIRY DATE: 02/04/2019  
VIOLATION CODE: DUI  
ARREST TIME: 00:21  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:53

SUBJECT'S LAST NAME: HENAO

DATE/TIME OF TEST: 10/25/2016 01:24:22

---

Diagnostics Check		OK	01:17
Air Blank		0.000	01:18
Control Test		0.081	01:18
Air Blank		0.000	01:18
Subject Sample	# 1	0.185	01:19
Breath Volume		2.242	01:19
Air Blank		0.000	01:20
Air Blank		0.000	01:22
Subject Sample	# 2	0.173	01:22
Breath Volume		1.273	01:22
Air Blank		0.000	01:23
Control Test		0.080	01:23
Air Blank		0.000	01:24
Diagnostics Check		OK	01:24

---

Data Download Date/Time: 11/18/2016 16:11

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: OWEN,SUE,  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: LOPEZ,BRANDON  
ARRESTING OFFICER AGENCY: Jupiter PD  
LAST AGENCY INSPECTION: 10/14/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 279870  
DRY GAS EXPIRY DATE: 02/04/2019  
VIOLATION CODE: DUI  
ARREST TIME: 00:23  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:06

SUBJECT'S LAST NAME: GARCIA NOLASCO

DATE/TIME OF TEST: 10/29/2016 01:39:17

---

Diagnostics Check	OK	01:29
Air Blank	0.000	01:30
Control Test	0.081	01:30
Air Blank	0.000	01:31
Subject Sample # 1	0.111	01:33
Breath Volume	1.207	01:33
Air Blank	0.000	01:33
Air Blank	0.000	01:35
Subject Sample # 2	0.103	01:37
Breath Volume	1.226	01:37
Air Blank	0.000	01:38
Control Test	0.081	01:38
Air Blank	0.000	01:39
Diagnostics Check	OK	01:39

---

Data Download Date/Time: 11/18/2016 16:11

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: OWEN,SUE,  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: CROY,PAUL  
ARRESTING OFFICER AGENCY: FHP Troop L  
LAST AGENCY INSPECTION: 10/14/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 279870  
DRY GAS EXPIRY DATE: 02/04/2019  
VIOLATION CODE: DUI  
ARREST TIME: 00:41  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:01

SUBJECT'S LAST NAME: LOPEZ-CAC

DATE/TIME OF TEST: 10/29/2016 02:32:12

---

Diagnostics Check	OK	02:25
Air Blank	0.000	02:26
Control Test	0.080	02:26
Air Blank	0.000	02:26
Subject Sample # 1	0.117	02:27
Breath Volume	1.917	02:27
Air Blank	0.000	02:28
Air Blank	0.000	02:30
Subject Sample # 2	0.120	02:30
Breath Volume	2.105	02:30
Air Blank	0.000	02:31
Control Test	0.081	02:31
Air Blank	0.000	02:32
Diagnostics Check	OK	02:32

---

Data Download Date/Time: 11/18/2016 16:11

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SCHAFFER,NANCY,M  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: MALLEY,C  
ARRESTING OFFICER AGENCY: Boynton Beach PD  
LAST AGENCY INSPECTION: 10/14/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 279870  
DRY GAS EXPIRY DATE: 02/04/2019  
VIOLATION CODE: DUI  
ARREST TIME: 18:22  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 18:55

SUBJECT'S LAST NAME: MARCOUX

DATE/TIME OF TEST: 10/29/2016 19:45:41

---

Diagnostics Check	OK	19:36
Air Blank	0.000	19:36
Control Test	0.081	19:37
Air Blank	0.000	19:37
Subject Sample # 1	0.000	19:40
Breath Volume	1.437	19:40
Air Blank	0.000	19:40
Air Blank	0.000	19:42
Subject Sample # 2	0.000	19:44
Breath Volume	1.304	19:44
Air Blank	0.000	19:44
Control Test	0.082	19:45
Air Blank	0.000	19:45
Diagnostics Check	OK	19:45

---

Data Download Date/Time: 11/18/2016 16:11

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: OWEN,SUE,  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: SORIANO,REYNALDO  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 10/14/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 279870  
DRY GAS EXPIRY DATE: 02/04/2019  
VIOLATION CODE: DUI  
ARREST TIME: 01:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: BONILLA-TOLEDO

DATE/TIME OF TEST: 10/30/2016 02:57:29

---

Diagnostics Check		OK	02:50
Air Blank		0.000	02:51
Control Test		0.081	02:51
Air Blank		0.000	02:52
Subject Sample	# 1	0.201	02:52
Breath Volume		2.296	02:52
Air Blank		0.000	02:53
Air Blank		0.000	02:55
Subject Sample	# 2	0.202	02:55
Breath Volume		1.335	02:55
Air Blank		0.000	02:56
Control Test		0.081	02:56
Air Blank		0.000	02:57
Diagnostics Check		OK	02:57

---

Data Download Date/Time: 11/18/2016 16:11

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: OWEN,SUE,  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: CHIQUITO  
RODRIGUEZ,LEENDY  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 10/14/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 279870  
DRY GAS EXPIRY DATE: 02/04/2019  
VIOLATION CODE: DUI  
ARREST TIME: 02:11  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: MENJIVAR

DATE/TIME OF TEST: 10/30/2016 03:50:56

---

Diagnostics Check		OK	03:38	
Air Blank		0.000	03:39	
Control Test		0.082	03:39	
Air Blank		0.000	03:39	
Subject Sample	# 1	0.126	03:43	Volume Not Met
Breath Volume		0.863	03:43	
Air Blank		0.000	03:43	
Air Blank		0.000	03:45	
Subject Sample	# 2	0.142	03:46	
Breath Volume		2.429	03:46	
Air Blank		0.000	03:46	
Air Blank		0.000	03:48	
Subject Sample	# 3	0.137	03:49	
Breath Volume		1.984	03:49	
Air Blank		0.000	03:50	
Control Test		0.080	03:50	
Air Blank		0.000	03:50	
Diagnostics Check		OK	03:50	

---

Data Download Date/Time: 11/18/2016 16:11

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PARENT,GARY,J  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: CHIQUITO-  
RODRIGUEZ,LENNDY  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 10/14/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 279870  
DRY GAS EXPIRY DATE: 02/04/2019  
VIOLATION CODE: DUI  
ARREST TIME: 21:49  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:11

SUBJECT'S LAST NAME: LEITHAUSER

DATE/TIME OF TEST: 10/31/2016 22:42:52

---

Diagnostics Check	OK	22:35
Air Blank	0.000	22:36
Control Test	0.081	22:36
Air Blank	0.000	22:37
Subject Sample # 1	0.369	22:37
Breath Volume	1.339	22:37
Air Blank	0.000	22:38
Air Blank	0.000	22:40
Subject Sample # 2	0.380	22:41
Breath Volume	1.707	22:41
Air Blank	0.000	22:41
Control Test	0.080	22:42
Air Blank	0.000	22:42
Diagnostics Check	OK	22:42

---

Data Download Date/Time: 11/18/2016 16:11

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: VANOVER,ANDREW,P  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: VANOVER,ANDREW  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 09/26/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1841081  
DRY GAS EXPIRY DATE: 04/08/2018  
VIOLATION CODE: DUI  
ARREST TIME: 22:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:32

SUBJECT'S LAST NAME: ANWAY

DATE/TIME OF TEST: 10/03/2016 23:59:13

---

Diagnostics Check	OK	23:53
Air Blank	0.000	23:53
Control Test	0.078	23:53
Air Blank	0.000	23:54
Subject Sample # 1	0.059	23:54
Breath Volume	1.652	23:54
Air Blank	0.000	23:55
Air Blank	0.000	23:57
Subject Sample # 2	0.056	23:57
Breath Volume	1.343	23:57
Air Blank	0.000	23:58
Control Test	0.077	23:58
Air Blank	0.000	23:59
Diagnostics Check	OK	23:59

---

Data Download Date/Time: 10/17/2016 10:45



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: AMBRIATI,NICHOLAS,J  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: AMBRIATI,NICHOLAS  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 09/26/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1841081  
DRY GAS EXPIRY DATE: 04/08/2018  
VIOLATION CODE: DUI  
ARREST TIME: 11:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 12:20

SUBJECT'S LAST NAME: SPORTS

DATE/TIME OF TEST: 10/06/2016 12:57:33

---

Diagnostics Check	OK	12:51
Air Blank	0.000	12:51
Control Test	0.079	12:52
Air Blank	0.000	12:52
Subject Sample # 1	0.000	12:53
Breath Volume	2.464	12:53
Air Blank	0.000	12:53
Air Blank	0.000	12:55
Subject Sample # 2	0.000	12:56
Breath Volume	1.652	12:56
Air Blank	0.000	12:56
Control Test	0.077	12:56
Air Blank	0.000	12:57
Diagnostics Check	OK	12:57

---

Data Download Date/Time: 10/17/2016 10:45

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: VANOVER,ANDREW,P  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: VANOVER,ANDREW  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 09/26/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1841081  
DRY GAS EXPIRY DATE: 04/08/2018  
VIOLATION CODE: DUI  
ARREST TIME: 21:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:55

SUBJECT'S LAST NAME: HERNANDEZ-SIERRA

DATE/TIME OF TEST: 10/07/2016 22:25:12

---

Diagnostics Check	OK	22:15
Air Blank	0.000	22:15
Control Test	0.078	22:16
Air Blank	0.000	22:16
Subject Sample # 1	0.204	22:17
Breath Volume	2.835	22:17
Air Blank	0.000	22:18
Air Blank	0.000	22:20
Subject Sample # 2	0.183	22:20
Breath Volume	1.476	22:20
Air Blank	0.000	22:21
Air Blank	0.000	22:23
Subject Sample # 3	0.189	22:23
Breath Volume	1.777	22:23
Air Blank	0.000	22:24
Control Test	0.075	22:24
Air Blank	0.000	22:25
Diagnostics Check	OK	22:25

---

Data Download Date/Time: 10/17/2016 10:45

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: VANOVER,ANDREW,P  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: VANOVER,ANDREW  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 09/26/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1841081  
DRY GAS EXPIRY DATE: 04/08/2018  
VIOLATION CODE: DUI  
ARREST TIME: 23:42  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:01

SUBJECT'S LAST NAME: GREEN

DATE/TIME OF TEST: 10/09/2016 00:36:06

---

Diagnostics Check	OK	00:33	
Air Blank	0.000	00:33	
Control Test	0.079	00:33	
Air Blank	0.000	00:34	
Subject Sample	# 1	00:34	Subject Test Refused
Breath Volume	0.000	00:34	
Air Blank	0.000	00:35	
Control Test	0.079	00:35	
Air Blank	0.000	00:36	
Diagnostics Check	OK	00:36	

---

Data Download Date/Time: 10/17/2016 10:45

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: VANOVER,ANDREW,P  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: VANOVER,ANDREW  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 09/26/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1841081  
DRY GAS EXPIRY DATE: 04/08/2018  
VIOLATION CODE: DUI  
ARREST TIME: 01:56  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:18

SUBJECT'S LAST NAME: BORJA-GOMEZ

DATE/TIME OF TEST: 10/09/2016 02:46:33

---

Diagnostics Check	OK	02:39
Air Blank	0.000	02:40
Control Test	0.078	02:40
Air Blank	0.000	02:41
Subject Sample # 1	0.138	02:42
Breath Volume	2.222	02:42
Air Blank	0.000	02:42
Air Blank	0.000	02:44
Subject Sample # 2	0.126	02:44
Breath Volume	1.753	02:44
Air Blank	0.000	02:45
Control Test	0.076	02:45
Air Blank	0.000	02:46
Diagnostics Check	OK	02:46

---

Data Download Date/Time: 10/17/2016 10:45

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: VANOVER,ANDREW,P  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: VANOVER,ANDREW  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 09/26/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1841081  
DRY GAS EXPIRY DATE: 04/08/2018  
VIOLATION CODE: DUI  
ARREST TIME: 01:10  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:34

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 10/13/2016 02:01:12

---

Diagnostics Check	OK	01:54
Air Blank	0.000	01:54
Control Test	0.078	01:55
Air Blank	0.000	01:55
Subject Sample # 1	0.328	01:56
Breath Volume	2.582	01:56
Air Blank	0.000	01:56
Air Blank	0.000	01:58
Subject Sample # 2	0.314	01:59
Breath Volume	1.476	01:59
Air Blank	0.000	02:00
Control Test	0.076	02:00
Air Blank	0.000	02:01
Diagnostics Check	OK	02:01

---

Data Download Date/Time: 10/17/2016 10:45

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BORES,ALAN,P  
BREATH TEST OPERATOR AGENCY: Holmes Beach PD  
ARRESTING OFFICER: BORES,ALAN  
ARRESTING OFFICER AGENCY: Holmes Beach PD  
LAST AGENCY INSPECTION: 09/26/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1841081  
DRY GAS EXPIRY DATE: 04/08/2018  
VIOLATION CODE: DUI  
ARREST TIME: 01:32  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:19

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 10/15/2016 02:43:33

---

Diagnostics Check	OK	02:40	
Air Blank	0.000	02:41	
Control Test	0.079	02:41	
Air Blank	0.000	02:41	
Subject Sample	# 1	02:42	Subject Test Refused
Breath Volume	0.000	02:42	
Air Blank	0.000	02:42	
Control Test	0.078	02:42	
Air Blank	0.000	02:43	
Diagnostics Check	OK	02:43	

---

Data Download Date/Time: 10/17/2016 10:45

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: VANOVER,ANDREW,P  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: VANOVER,ANDREW  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 10/17/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1841081  
DRY GAS EXPIRY DATE: 04/08/2018  
VIOLATION CODE: DUI  
ARREST TIME: 20:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:34

SUBJECT'S LAST NAME: FAY JR

DATE/TIME OF TEST: 10/18/2016 21:04:11

---

Diagnostics Check	OK	20:55
Air Blank	0.000	20:56
Control Test	0.078	20:56
Air Blank	0.000	20:57
Subject Sample # 1	0.000	20:58
Breath Volume	1.500	20:58
Air Blank	0.000	20:58
Air Blank	0.000	21:00
Subject Sample # 2	0.000	21:02
Breath Volume	1.406	21:02
Air Blank	0.000	21:03
Control Test	0.077	21:03
Air Blank	0.000	21:04
Diagnostics Check	OK	21:04

---

Data Download Date/Time: 11/28/2016 09:13

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: QUICK,BENJAMIN,J  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: QUICK,BENJAMIN  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 10/17/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1841081  
DRY GAS EXPIRY DATE: 04/08/2018  
VIOLATION CODE: DUI  
ARREST TIME: 00:24  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:57

SUBJECT'S LAST NAME: NIVER

DATE/TIME OF TEST: 10/22/2016 01:24:29

---

Diagnostics Check	OK	01:17
Air Blank	0.000	01:18
Control Test	0.079	01:18
Air Blank	0.000	01:19
Subject Sample # 1	0.185	01:19
Breath Volume	3.816	01:19
Air Blank	0.000	01:20
Air Blank	0.000	01:22
Subject Sample # 2	0.183	01:22
Breath Volume	3.425	01:22
Air Blank	0.000	01:23
Control Test	0.078	01:23
Air Blank	0.000	01:24
Diagnostics Check	OK	01:24

---

Data Download Date/Time: 11/28/2016 09:13



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: VANOVER,ANDREW,P  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: VANOVER,ANDREW  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 10/17/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1841081  
DRY GAS EXPIRY DATE: 04/08/2018  
VIOLATION CODE: DUI  
ARREST TIME: 02:23  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:44

SUBJECT'S LAST NAME: KEMP

DATE/TIME OF TEST: 10/22/2016 03:09:31

---

Diagnostics Check	OK	03:06	
Air Blank	0.000	03:06	
Control Test	0.079	03:07	
Air Blank	0.000	03:07	
Subject Sample	# 1	03:08	Subject Test Refused
Breath Volume	0.000	03:08	
Air Blank	0.000	03:08	
Control Test	0.079	03:08	
Air Blank	0.000	03:09	
Diagnostics Check	OK	03:09	

---

Data Download Date/Time: 11/28/2016 09:13

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: VANOVER,ANDREW,P  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: VANOVER,ANDREW  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 10/17/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1841081  
DRY GAS EXPIRY DATE: 04/08/2018  
VIOLATION CODE: DUI  
ARREST TIME: 19:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:20

SUBJECT'S LAST NAME: LANDKE

DATE/TIME OF TEST: 10/23/2016 20:47:11

---

Diagnostics Check	OK	20:40
Air Blank	0.000	20:41
Control Test	0.079	20:41
Air Blank	0.000	20:42
Subject Sample # 1	0.123	20:42
Breath Volume	2.597	20:42
Air Blank	0.000	20:43
Air Blank	0.000	20:45
Subject Sample # 2	0.118	20:45
Breath Volume	1.488	20:45
Air Blank	0.000	20:46
Control Test	0.077	20:46
Air Blank	0.000	20:47
Diagnostics Check	OK	20:47

---

Data Download Date/Time: 11/28/2016 09:13

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: VANOVER,ANDREW,P  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: VANOVER,ANDREW  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 10/17/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1841081  
DRY GAS EXPIRY DATE: 04/08/2018  
VIOLATION CODE: DUI  
ARREST TIME: 22:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 10/24/2016 00:26:12

---

Diagnostics Check	OK	00:23	
Air Blank	0.000	00:23	
Control Test	0.080	00:24	
Air Blank	0.000	00:24	
Subject Sample	# 1	00:24	Subject Test Refused
Breath Volume	0.000	00:24	
Air Blank	0.000	00:25	
Control Test	0.079	00:25	
Air Blank	0.000	00:26	
Diagnostics Check	OK	00:26	

---

Data Download Date/Time: 11/28/2016 09:13

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: VANOVER,ANDREW,P  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: VANOVER,ANDREW  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 10/17/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1841081  
DRY GAS EXPIRY DATE: 04/08/2018  
VIOLATION CODE: DUI  
ARREST TIME: 00:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: PAULDEN

DATE/TIME OF TEST: 10/24/2016 01:08:28

---

Diagnostics Check	OK	01:05	
Air Blank	0.000	01:05	
Control Test	0.079	01:05	
Air Blank	0.000	01:06	
Subject Sample	# 1	01:07	Subject Test Refused
Breath Volume	0.000	01:07	
Air Blank	0.000	01:07	
Control Test	0.079	01:07	
Air Blank	0.000	01:08	
Diagnostics Check	OK	01:08	

---

Data Download Date/Time: 11/28/2016 09:13

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KENYAN, MATTHEW, A  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: KENYAN, M  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 10/17/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1841081  
DRY GAS EXPIRY DATE: 04/08/2018  
VIOLATION CODE: DUI  
ARREST TIME: 04:14  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:12

SUBJECT'S LAST NAME: FORDHAM

DATE/TIME OF TEST: 10/26/2016 05:39:17

---

Diagnostics Check	OK	05:32
Air Blank	0.000	05:33
Control Test	0.079	05:33
Air Blank	0.000	05:34
Subject Sample # 1	0.083	05:34
Breath Volume	3.433	05:34
Air Blank	0.000	05:35
Air Blank	0.000	05:37
Subject Sample # 2	0.080	05:37
Breath Volume	1.898	05:37
Air Blank	0.000	05:38
Control Test	0.078	05:38
Air Blank	0.000	05:39
Diagnostics Check	OK	05:39

---

Data Download Date/Time: 11/28/2016 09:13

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: QUICK,BENJAMIN,J  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: QUICK,BENJAMIN  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 10/17/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1841081  
DRY GAS EXPIRY DATE: 04/08/2018  
VIOLATION CODE: DUI  
ARREST TIME: 22:02  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:58

SUBJECT'S LAST NAME: HICKS

DATE/TIME OF TEST: 10/26/2016 23:26:17

---

Diagnostics Check		OK	23:19
Air Blank		0.000	23:19
Control Test		0.078	23:19
Air Blank		0.000	23:20
Subject Sample	# 1	0.204	23:20
Breath Volume		2.421	23:20
Air Blank		0.000	23:21
Air Blank		0.000	23:23
Subject Sample	# 2	0.199	23:24
Breath Volume		2.476	23:24
Air Blank		0.000	23:25
Control Test		0.077	23:25
Air Blank		0.000	23:26
Diagnostics Check		OK	23:26

---

Data Download Date/Time: 11/28/2016 09:13

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TATE,ROBERT,J  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: TATE,ROBERT  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 09/26/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 256835  
DRY GAS EXPIRY DATE: 01/11/2019  
VIOLATION CODE: DUI  
ARREST TIME: 17:13  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:40

SUBJECT'S LAST NAME: ELWOOD

DATE/TIME OF TEST: 10/02/2016 19:08:44

---

Diagnostics Check	OK	19:00
Air Blank	0.000	19:01
Control Test	0.077	19:01
Air Blank	0.000	19:02
Subject Sample # 1	0.000	19:03
Breath Volume	2.011	19:03
Air Blank	0.000	19:04
Air Blank	0.000	19:06
Subject Sample # 2	0.000	19:07
Breath Volume	1.578	19:07
Air Blank	0.000	19:07
Control Test	0.078	19:08
Air Blank	0.000	19:08
Diagnostics Check	OK	19:08

---

Data Download Date/Time: 10/17/2016 12:39

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROWE,BRANDON,D  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: ROWE,BRANDON  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 09/26/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 256835  
DRY GAS EXPIRY DATE: 01/11/2019  
VIOLATION CODE: DUI  
ARREST TIME: 22:01  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: GIORLANDO

DATE/TIME OF TEST: 10/03/2016 23:07:39

---

Diagnostics Check	OK	23:01
Air Blank	0.000	23:01
Control Test	0.079	23:01
Air Blank	0.000	23:02
Subject Sample # 1	0.138	23:02
Breath Volume	1.832	23:02
Air Blank	0.000	23:03
Air Blank	0.000	23:05
Subject Sample # 2	0.126	23:06
Breath Volume	1.621	23:06
Air Blank	0.000	23:06
Control Test	0.078	23:07
Air Blank	0.000	23:07
Diagnostics Check	OK	23:07

---

Data Download Date/Time: 10/17/2016 12:39



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: OSTERMEYER,DAMIAN,J  
BREATH TEST OPERATOR AGENCY: FHP Troop F  
ARRESTING OFFICER: OSTERMEYER,DAMIAN  
ARRESTING OFFICER AGENCY: FHP Troop F  
LAST AGENCY INSPECTION: 09/26/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 256835  
DRY GAS EXPIRY DATE: 01/11/2019  
VIOLATION CODE: DUI  
ARREST TIME: 13:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 14:05

SUBJECT'S LAST NAME: SPONG

DATE/TIME OF TEST: 10/04/2016 15:00:20

---

Diagnostics Check	OK	14:59	
Air Blank		14:59	Ambient Fail
Air Blank	0.000	15:00	

---

Data Download Date/Time: 10/17/2016 12:39

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: OSTERMEYER,DAMIAN,J  
BREATH TEST OPERATOR AGENCY: FHP Troop F  
ARRESTING OFFICER: OSTERMEYE,DAMIAN  
ARRESTING OFFICER AGENCY: FHP Troop F  
LAST AGENCY INSPECTION: 09/26/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 256835  
DRY GAS EXPIRY DATE: 01/11/2019  
VIOLATION CODE: DUI  
ARREST TIME: 13:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 14:05

SUBJECT'S LAST NAME: SPONG

DATE/TIME OF TEST: 10/04/2016 15:14:17

---

Diagnostics Check	OK	15:03
Air Blank	0.000	15:03
Control Test	0.079	15:03
Air Blank	0.000	15:04
Subject Sample # 1	0.079	15:07
Breath Volume	1.769	15:07
Air Blank	0.000	15:07
Air Blank	0.000	15:09
Subject Sample # 2	0.071	15:12
Breath Volume	1.449	15:12
Air Blank	0.000	15:13
Control Test	0.077	15:13
Air Blank	0.000	15:14
Diagnostics Check	OK	15:14

---

Data Download Date/Time: 10/17/2016 12:39

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: VANOVER,ANDREW,P  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: VANOVER,ANDREW  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 09/26/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 256835  
DRY GAS EXPIRY DATE: 01/11/2019  
VIOLATION CODE: DUI  
ARREST TIME: 22:32  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:12

SUBJECT'S LAST NAME: WOODALL

DATE/TIME OF TEST: 10/04/2016 23:39:06

---

Diagnostics Check	OK	23:32
Air Blank	0.000	23:33
Control Test	0.079	23:33
Air Blank	0.000	23:34
Subject Sample # 1	0.332	23:34
Breath Volume	2.402	23:34
Air Blank	0.000	23:35
Air Blank	0.000	23:36
Subject Sample # 2	0.331	23:37
Breath Volume	2.390	23:37
Air Blank	0.000	23:38
Control Test	0.076	23:38
Air Blank	0.000	23:39
Diagnostics Check	OK	23:39

---

Data Download Date/Time: 10/17/2016 12:39

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PETERSON,STEFANIE,M  
BREATH TEST OPERATOR AGENCY: FHP Troop F  
ARRESTING OFFICER: PETERSON,STEFANIE  
ARRESTING OFFICER AGENCY: FHP Troop F  
LAST AGENCY INSPECTION: 09/26/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 256835  
DRY GAS EXPIRY DATE: 01/11/2019  
VIOLATION CODE: DUI  
ARREST TIME: 23:57  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: HART

DATE/TIME OF TEST: 10/07/2016 01:27:08

---

Diagnostics Check	OK	01:20
Air Blank	0.000	01:21
Control Test	0.078	01:21
Air Blank	0.000	01:22
Subject Sample # 1	0.221	01:22
Breath Volume	1.613	01:22
Air Blank	0.000	01:23
Air Blank	0.000	01:25
Subject Sample # 2	0.205	01:25
Breath Volume	1.429	01:25
Air Blank	0.000	01:26
Control Test	0.077	01:26
Air Blank	0.000	01:27
Diagnostics Check	OK	01:27

---

Data Download Date/Time: 10/17/2016 12:39

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SAYED, JEREMIAH, A  
BREATH TEST OPERATOR AGENCY: FHP Troop F  
ARRESTING OFFICER: SAYED, JEREMIAH  
ARRESTING OFFICER AGENCY: FHP Troop F  
LAST AGENCY INSPECTION: 09/26/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 256835  
DRY GAS EXPIRY DATE: 01/11/2019  
VIOLATION CODE: DUI  
ARREST TIME: 05:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:15

SUBJECT'S LAST NAME: HOLDER

DATE/TIME OF TEST: 10/08/2016 07:17:05

---

Diagnostics Check	OK	07:10
Air Blank	0.000	07:10
Control Test	0.079	07:11
Air Blank	0.000	07:11
Subject Sample # 1	0.015	07:12
Breath Volume	1.554	07:12
Air Blank	0.000	07:12
Air Blank	0.000	07:14
Subject Sample # 2	0.015	07:15
Breath Volume	1.613	07:15
Air Blank	0.000	07:16
Control Test	0.079	07:16
Air Blank	0.000	07:17
Diagnostics Check	OK	07:17

---

Data Download Date/Time: 10/17/2016 12:39

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PAIGE,KYLE,C  
BREATH TEST OPERATOR AGENCY: FHP Troop F  
ARRESTING OFFICER: PAIGE,KYLE  
ARRESTING OFFICER AGENCY: FHP Troop F  
LAST AGENCY INSPECTION: 09/26/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 256835  
DRY GAS EXPIRY DATE: 01/11/2019  
VIOLATION CODE: DUI  
ARREST TIME: 10:19  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 11:18

SUBJECT'S LAST NAME: BARNETT

DATE/TIME OF TEST: 10/10/2016 11:47:13

---

Diagnostics Check	OK	11:40
Air Blank	0.000	11:41
Control Test	0.079	11:41
Air Blank	0.000	11:42
Subject Sample # 1	0.000	11:42
Breath Volume	1.265	11:42
Air Blank	0.000	11:43
Air Blank	0.000	11:45
Subject Sample # 2	0.000	11:45
Breath Volume	1.195	11:45
Air Blank	0.000	11:46
Control Test	0.077	11:46
Air Blank	0.000	11:47
Diagnostics Check	OK	11:47

---

Data Download Date/Time: 10/17/2016 12:39

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TATE,ROBERT,J  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: TATE,ROBERT  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 09/26/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 256835  
DRY GAS EXPIRY DATE: 01/11/2019  
VIOLATION CODE: DUI  
ARREST TIME: 00:17  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: BIGLER

DATE/TIME OF TEST: 10/15/2016 01:52:52

---

Diagnostics Check		OK	01:41	
Air Blank		0.000	01:41	
Control Test		0.077	01:41	
Air Blank		0.000	01:42	
Subject Sample	# 1	0.196	01:45	Volume Not Met
Breath Volume		0.589	01:45	
Air Blank		0.000	01:46	
Air Blank		0.000	01:47	
Subject Sample	# 2	0.181	01:51	Volume Not Met
Breath Volume		0.097	01:51	
Air Blank		0.000	01:51	
Control Test		0.076	01:52	
Air Blank		0.000	01:52	
Diagnostics Check		OK	01:52	

---

Data Download Date/Time: 10/17/2016 12:39

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: OSTERMEYER,DAMIAN,J  
BREATH TEST OPERATOR AGENCY: FHP Troop F  
ARRESTING OFFICER: OSTERMEYER,DAMIAN  
ARRESTING OFFICER AGENCY: FHP Troop F  
LAST AGENCY INSPECTION: 09/26/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 256835  
DRY GAS EXPIRY DATE: 01/11/2019  
VIOLATION CODE: DUI  
ARREST TIME: 11:39  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:04

SUBJECT'S LAST NAME: SUDOL

DATE/TIME OF TEST: 10/16/2016 01:28:26

---

Diagnostics Check	OK	01:25	
Air Blank	0.000	01:25	
Control Test	0.079	01:26	
Air Blank	0.000	01:26	
Subject Sample	# 1	01:27	Subject Test Refused
Breath Volume	0.000	01:27	
Air Blank	0.000	01:27	
Control Test	0.079	01:27	
Air Blank	0.000	01:28	
Diagnostics Check	OK	01:28	

---

Data Download Date/Time: 10/17/2016 12:39



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: N,TAYLOR,M  
BREATH TEST OPERATOR AGENCY: FHP Troop F  
ARRESTING OFFICER: LEDFORD,TAYLOR  
ARRESTING OFFICER AGENCY: FHP Troop F  
LAST AGENCY INSPECTION: 10/17/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 256835  
DRY GAS EXPIRY DATE: 01/11/2019  
VIOLATION CODE: DUI  
ARREST TIME: 21:48  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:28

SUBJECT'S LAST NAME: LEDFORD

DATE/TIME OF TEST: 10/20/2016 22:55:53

---

Diagnostics Check	OK	22:49
Air Blank	0.000	22:49
Control Test	0.079	22:49
Air Blank	0.000	22:50
Subject Sample # 1	0.028	22:51
Breath Volume	1.964	22:51
Air Blank	0.000	22:52
Air Blank	0.000	22:53
Subject Sample # 2	0.025	22:54
Breath Volume	1.324	22:54
Air Blank	0.000	22:54
Control Test	0.077	22:55
Air Blank	0.000	22:55
Diagnostics Check	OK	22:55

---

Data Download Date/Time: 11/28/2016 09:16

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TATE,ROBERT,J  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: TATE,ROBERT  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 10/17/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 256835  
DRY GAS EXPIRY DATE: 01/11/2019  
VIOLATION CODE: DUI  
ARREST TIME: 23:02  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: MAYA

DATE/TIME OF TEST: 10/21/2016 00:18:44

---

Diagnostics Check	OK	00:07	
Air Blank	0.000	00:07	
Control Test	0.079	00:08	
Air Blank	0.000	00:08	
Subject Sample	# 1	00:11	No Sample Provided
Breath Volume	0.000	00:11	
Air Blank	0.000	00:12	
Air Blank	0.000	00:14	
Subject Sample	# 2	00:17	No Sample Provided
Breath Volume	0.000	00:17	
Air Blank	0.000	00:17	
Control Test	0.079	00:18	
Air Blank	0.000	00:18	
Diagnostics Check	OK	00:18	

---

Data Download Date/Time: 11/28/2016 09:16

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: OSTERMEYER,DAMIAN,J  
BREATH TEST OPERATOR AGENCY: FHP Troop F  
ARRESTING OFFICER: OSTERMEYER,DAMIAN  
ARRESTING OFFICER AGENCY: FHP Troop F  
LAST AGENCY INSPECTION: 10/17/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 256835  
DRY GAS EXPIRY DATE: 01/11/2019  
VIOLATION CODE: DUI  
ARREST TIME: 04:41  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:06

SUBJECT'S LAST NAME: SCHWEITZER

DATE/TIME OF TEST: 10/21/2016 09:57:15

---

Diagnostics Check	OK	09:56	
Air Blank		09:56	RFI Detect
Air Blank		09:57	RFI Detect

---

Data Download Date/Time: 11/28/2016 09:16

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: OSTERMEYER,DAMIAN,J  
BREATH TEST OPERATOR AGENCY: FHP Troop F  
ARRESTING OFFICER: OSTERMEYE,DAMIAN  
ARRESTING OFFICER AGENCY: FHP Troop F  
LAST AGENCY INSPECTION: 10/17/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 256835  
DRY GAS EXPIRY DATE: 01/11/2019  
VIOLATION CODE: DUI  
ARREST TIME: 04:41  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:06

SUBJECT'S LAST NAME: SCHWEITZER

DATE/TIME OF TEST: 10/21/2016 10:02:30

---

Diagnostics Check	OK	09:59	
Air Blank	0.000	10:00	
Control Test	0.079	10:00	
Air Blank	0.000	10:00	
Subject Sample	# 1	10:01	Subject Test Refused
Breath Volume	0.000	10:01	
Air Blank	0.000	10:01	
Control Test	0.079	10:01	
Air Blank	0.000	10:02	
Diagnostics Check	OK	10:02	

---

Data Download Date/Time: 11/28/2016 09:16

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: STANLEY,JEFFREY,A  
BREATH TEST OPERATOR AGENCY: FHP Troop F  
ARRESTING OFFICER: STANLEY,JEFFREY  
ARRESTING OFFICER AGENCY: FHP Troop F  
LAST AGENCY INSPECTION: 10/17/2016

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 256835  
DRY GAS EXPIRY DATE: 01/11/2019  
VIOLATION CODE: DUI  
ARREST TIME: 19:27  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:30

SUBJECT'S LAST NAME: SMYTH

DATE/TIME OF TEST: 10/26/2016 20:58:44

---

Diagnostics Check	OK	20:52	
Air Blank	0.000	20:52	
Control Test	0.079	20:53	
Air Blank	0.000	20:53	
Subject Sample	# 1 0.202	20:54	
Breath Volume	2.578	20:54	
Air Blank	0.000	20:55	
Air Blank	0.000	20:56	
Subject Sample	# 2 0.204	20:57	
Breath Volume	2.937	20:57	
Air Blank		20:58	RFI Detect
Air Blank	0.000	20:58	

---

Data Download Date/Time: 11/28/2016 09:17

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TATE,ROBERT,J  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: TATE,ROBERT  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 10/17/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 256835  
DRY GAS EXPIRY DATE: 01/11/2019  
VIOLATION CODE: DUI  
ARREST TIME: 02:56  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: BORJAS-GAITANI

DATE/TIME OF TEST: 10/29/2016 03:51:54

---

Diagnostics Check	OK	03:45
Air Blank	0.000	03:46
Control Test	0.079	03:46
Air Blank	0.000	03:46
Subject Sample # 1	0.176	03:47
Breath Volume	1.246	03:47
Air Blank	0.000	03:48
Air Blank	0.000	03:49
Subject Sample # 2	0.194	03:50
Breath Volume	1.398	03:50
Air Blank	0.000	03:50
Control Test	0.078	03:51
Air Blank	0.000	03:51
Diagnostics Check	OK	03:51

---

Data Download Date/Time: 11/28/2016 09:17

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DIAZ,LUCCIANO,R  
BREATH TEST OPERATOR AGENCY: Palmetto PD  
ARRESTING OFFICER: VILA-REED,J  
ARRESTING OFFICER AGENCY: Palmetto PD  
LAST AGENCY INSPECTION: 10/17/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 256835  
DRY GAS EXPIRY DATE: 01/11/2019  
VIOLATION CODE: DUI  
ARREST TIME: 12:43  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 14:05

SUBJECT'S LAST NAME: MENDEZ-CALMO

DATE/TIME OF TEST: 10/29/2016 14:43:05

---

Diagnostics Check	OK	14:27	
Air Blank	0.000	14:27	
Control Test	0.079	14:28	
Air Blank	0.000	14:28	
Subject Sample	# 1 0.155	14:30	
Breath Volume	1.453	14:30	
Air Blank	0.000	14:31	
Air Blank	0.000	14:32	
Subject Sample	# 2 0.163	14:36	Volume Not Met
Breath Volume	0.031	14:36	
Air Blank	0.000	14:36	
Air Blank	0.000	14:38	
Subject Sample	# 3	14:41	No Sample Provided
Breath Volume	0.000	14:41	
Air Blank	0.000	14:42	
Control Test	0.076	14:42	
Air Blank	0.000	14:43	
Diagnostics Check	OK	14:43	

---

Data Download Date/Time: 11/28/2016 09:17

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TATE,ROBERT,J  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: TATE,ROBERT  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 10/17/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 256835  
DRY GAS EXPIRY DATE: 01/11/2019  
VIOLATION CODE: DUI  
ARREST TIME: 01:38  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: BELLO

DATE/TIME OF TEST: 10/30/2016 02:43:04

---

Diagnostics Check	OK	02:37	
Air Blank	0.000	02:37	
Control Test	0.078	02:38	
Air Blank	0.000	02:38	
Subject Sample	# 1	02:41	Subject Test Refused
Breath Volume	0.000	02:41	
Air Blank	0.000	02:42	
Control Test	0.078	02:42	
Air Blank	0.000	02:42	
Diagnostics Check	OK	02:43	

---

Data Download Date/Time: 11/28/2016 09:17



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROWE,BRANDON,D  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: JONES,JOHN  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 10/17/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 256835  
DRY GAS EXPIRY DATE: 01/11/2019  
VIOLATION CODE: DUI  
ARREST TIME: 02:47  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 10/30/2016 04:29:03

---

Diagnostics Check		OK	04:22
Air Blank		0.000	04:23
Control Test		0.079	04:23
Air Blank		0.000	04:24
Subject Sample	# 1	0.162	04:24
Breath Volume		2.207	04:24
Air Blank		0.000	04:25
Air Blank		0.000	04:27
Subject Sample	# 2	0.159	04:27
Breath Volume		2.683	04:27
Air Blank		0.000	04:28
Control Test		0.077	04:28
Air Blank		0.000	04:28
Diagnostics Check		OK	04:29

---

Data Download Date/Time: 11/28/2016 09:17

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BORES,ALAN,P  
BREATH TEST OPERATOR AGENCY: Holmes Beach PD  
ARRESTING OFFICER: BORES,ALAN  
ARRESTING OFFICER AGENCY: Holmes Beach PD  
LAST AGENCY INSPECTION: 10/17/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 256835  
DRY GAS EXPIRY DATE: 01/11/2019  
VIOLATION CODE: DUI  
ARREST TIME: 03:08  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: PASIK

DATE/TIME OF TEST: 10/30/2016 04:39:49

---

Diagnostics Check	OK	04:33
Air Blank	0.000	04:33
Control Test	0.077	04:34
Air Blank	0.000	04:34
Subject Sample # 1	0.121	04:35
Breath Volume	2.507	04:35
Air Blank	0.000	04:35
Air Blank	0.000	04:37
Subject Sample # 2	0.119	04:38
Breath Volume	1.894	04:38
Air Blank	0.000	04:38
Control Test	0.078	04:39
Air Blank	0.000	04:39
Diagnostics Check	OK	04:39

---

Data Download Date/Time: 11/28/2016 09:17

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MULLIGAN,WILLIAM,B  
BREATH TEST OPERATOR AGENCY: Bradenton PD  
ARRESTING OFFICER: SANCHEZ,JOSHUA  
ARRESTING OFFICER AGENCY: Bradenton PD  
LAST AGENCY INSPECTION: 10/17/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 256835  
DRY GAS EXPIRY DATE: 01/11/2019  
VIOLATION CODE: DUI  
ARREST TIME: 03:39  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:57

SUBJECT'S LAST NAME: MOORE

DATE/TIME OF TEST: 10/30/2016 05:24:49

---

Diagnostics Check	OK	05:21	
Air Blank	0.000	05:22	
Control Test	0.078	05:22	
Air Blank	0.000	05:22	
Subject Sample	# 1	05:23	Subject Test Refused
Breath Volume	0.000	05:23	
Air Blank	0.000	05:23	
Control Test	0.079	05:24	
Air Blank	0.000	05:24	
Diagnostics Check	OK	05:24	

---

Data Download Date/Time: 11/28/2016 09:17

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M  
BREATH TEST OPERATOR AGENCY: FHP Troop F  
ARRESTING OFFICER: LEDFORD,TAYLOR  
ARRESTING OFFICER AGENCY: FHP Troop F  
LAST AGENCY INSPECTION: 10/17/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 256835  
DRY GAS EXPIRY DATE: 01/11/2019  
VIOLATION CODE: DUI  
ARREST TIME: 00:38  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: BAUTISTA HERNANDEZ

DATE/TIME OF TEST: 10/31/2016 01:45:30

---

Diagnostics Check	OK	01:36
Air Blank	0.000	01:36
Control Test	0.079	01:36
Air Blank	0.000	01:37
Subject Sample # 1	0.251	01:40
Breath Volume	1.953	01:40
Air Blank	0.000	01:41
Air Blank	0.000	01:43
Subject Sample # 2	0.243	01:43
Breath Volume	1.523	01:43
Air Blank	0.000	01:44
Control Test	0.075	01:44
Air Blank	0.000	01:45
Diagnostics Check	OK	01:45

---

Data Download Date/Time: 11/28/2016 09:17

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BRENNAN,D  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 09/07/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 12716080A7  
DRY GAS EXPIRY DATE: 06/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 04:40  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 05:15

SUBJECT'S LAST NAME: MURILLO

DATE/TIME OF TEST: 10/01/2016 05:42:31

---

Diagnostics Check	OK	05:36
Air Blank	0.000	05:36
Control Test	0.079	05:36
Air Blank	0.000	05:37
Subject Sample # 1	0.167	05:37
Breath Volume	2.589	05:37
Air Blank	0.000	05:38
Air Blank	0.000	05:40
Subject Sample # 2	0.163	05:40
Breath Volume	2.289	05:40
Air Blank	0.000	05:41
Control Test	0.078	05:41
Air Blank	0.000	05:42
Diagnostics Check	OK	05:42

---

Data Download Date/Time: 10/05/2016 17:42

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SIMMONS,ANGELA,V  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: HAZELZET,DAVID  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 09/07/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 12716080A7  
DRY GAS EXPIRY DATE: 06/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 06:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 07:28

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 10/01/2016 07:55:18

---

Diagnostics Check	OK	07:48
Air Blank	0.000	07:49
Control Test	0.080	07:49
Air Blank	0.000	07:50
Subject Sample # 1	0.117	07:50
Breath Volume	2.070	07:50
Air Blank	0.000	07:51
Air Blank	0.000	07:53
Subject Sample # 2	0.117	07:53
Breath Volume	1.480	07:53
Air Blank	0.000	07:54
Control Test	0.078	07:54
Air Blank	0.000	07:55
Diagnostics Check	OK	07:55

---

Data Download Date/Time: 10/05/2016 17:42

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HENRY,PATRICIA,  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: FOWLER,J  
ARRESTING OFFICER AGENCY: Plant City PD  
LAST AGENCY INSPECTION: 09/07/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 12716080A7  
DRY GAS EXPIRY DATE: 06/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 18:36  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:16

SUBJECT'S LAST NAME: BALL

DATE/TIME OF TEST: 10/01/2016 19:42:33

---

Diagnostics Check	OK	19:36
Air Blank	0.000	19:36
Control Test	0.079	19:36
Air Blank	0.000	19:37
Subject Sample # 1	0.297	19:37
Breath Volume	3.015	19:37
Air Blank	0.000	19:38
Air Blank	0.000	19:40
Subject Sample # 2	0.287	19:40
Breath Volume	3.324	19:40
Air Blank	0.000	19:41
Control Test	0.078	19:41
Air Blank	0.000	19:42
Diagnostics Check	OK	19:42

---

Data Download Date/Time: 10/05/2016 17:42

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TORRES,JENIFER,  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BRABAND,M  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 09/07/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 12716080A7  
DRY GAS EXPIRY DATE: 06/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 21:42  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:19

SUBJECT'S LAST NAME: SANTIAGO-CHAVEZ

DATE/TIME OF TEST: 10/01/2016 22:47:01

---

Diagnostics Check	OK	22:39
Air Blank	0.000	22:39
Control Test	0.079	22:40
Air Blank	0.000	22:40
Subject Sample # 1	0.159	22:41
Breath Volume	1.871	22:41
Air Blank	0.000	22:42
Air Blank	0.000	22:43
Subject Sample # 2	0.166	22:45
Breath Volume	2.312	22:45
Air Blank	0.000	22:46
Control Test	0.079	22:46
Air Blank	0.000	22:46
Diagnostics Check	OK	22:47

---

Data Download Date/Time: 10/05/2016 17:42



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TRIMBLE, TERESA,  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: GONZALEZ, D  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 09/07/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 12716080A7  
DRY GAS EXPIRY DATE: 06/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 04:42  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:03

SUBJECT'S LAST NAME: DAWSON

DATE/TIME OF TEST: 10/02/2016 05:38:57

---

Diagnostics Check		OK	05:24	
Air Blank		0.000	05:24	
Control Test		0.079	05:25	
Air Blank		0.000	05:25	
Subject Sample	# 1	0.210	05:27	No .020 agreement
Breath Volume		2.148	05:27	
Air Blank		0.000	05:28	
Air Blank		0.000	05:29	
Subject Sample	# 2	0.203	05:33	Volume Not Met
Breath Volume		0.074	05:33	
Air Blank		0.000	05:33	
Air Blank		0.000	05:35	
Subject Sample	# 3	0.180	05:37	No .020 agreement
Breath Volume		2.253	05:37	
Air Blank		0.000	05:38	
Control Test		0.076	05:38	
Air Blank		0.000	05:38	
Diagnostics Check		OK	05:38	

---

Data Download Date/Time: 10/05/2016 17:42

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TRIMBLE, TERESA,  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: GONZALEZ, D  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 09/07/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 12716080A7  
DRY GAS EXPIRY DATE: 06/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 04:42  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:03

SUBJECT'S LAST NAME: DAWSON

DATE/TIME OF TEST: 10/02/2016 05:52:20

---

Diagnostics Check	OK	05:41
Air Blank	0.000	05:42
Control Test	0.078	05:42
Air Blank	0.000	05:43
Subject Sample # 1	0.205	05:45
Breath Volume	1.394	05:45
Air Blank	0.000	05:45
Air Blank	0.000	05:47
Subject Sample # 2	0.196	05:50
Breath Volume	2.074	05:50
Air Blank	0.000	05:51
Control Test	0.076	05:51
Air Blank	0.000	05:52
Diagnostics Check	OK	05:52

---

Data Download Date/Time: 10/05/2016 17:42

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: VAN-TREESE,S  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 09/07/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 12716080A7  
DRY GAS EXPIRY DATE: 06/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 02:21  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:46

SUBJECT'S LAST NAME: BOGERTY JR

DATE/TIME OF TEST: 10/05/2016 03:18:24

---

Diagnostics Check		OK	03:06	
Air Blank		0.000	03:07	
Control Test		0.080	03:07	
Air Blank		0.000	03:08	
Subject Sample	# 1	0.190	03:11	Volume Not Met
Breath Volume		0.234	03:11	
Air Blank		0.000	03:11	
Air Blank		0.000	03:13	
Subject Sample	# 2	0.145	03:16	Volume Not Met
Breath Volume		0.257	03:16	
Air Blank		0.000	03:17	
Control Test		0.078	03:17	
Air Blank		0.000	03:18	
Diagnostics Check		OK	03:18	

---

Data Download Date/Time: 10/05/2016 17:42

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KIRK,CASEY,S  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BRENNAN,D  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 10/05/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 12716080A7  
DRY GAS EXPIRY DATE: 06/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 23:52  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: MCGUIRE

DATE/TIME OF TEST: 10/06/2016 00:57:10

---

Diagnostics Check		OK	00:50
Air Blank		0.000	00:51
Control Test		0.080	00:51
Air Blank		0.000	00:52
Subject Sample	# 1	0.259	00:52
Breath Volume		2.644	00:52
Air Blank		0.000	00:53
Air Blank		0.000	00:54
Subject Sample	# 2	0.259	00:55
Breath Volume		2.621	00:55
Air Blank		0.000	00:56
Control Test		0.079	00:56
Air Blank		0.000	00:57
Diagnostics Check		OK	00:57

---

Data Download Date/Time: 11/07/2016 20:33

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MCMINN,KATHY,  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BRENNAN,D  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 10/05/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 12716080A7  
DRY GAS EXPIRY DATE: 06/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 05:41  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:08

SUBJECT'S LAST NAME: KNIGHT

DATE/TIME OF TEST: 10/06/2016 06:34:54

---

Diagnostics Check	OK	06:28
Air Blank	0.000	06:28
Control Test	0.080	06:29
Air Blank	0.000	06:29
Subject Sample # 1	0.079	06:30
Breath Volume	3.683	06:30
Air Blank	0.000	06:31
Air Blank	0.000	06:32
Subject Sample # 2	0.078	06:33
Breath Volume	1.328	06:33
Air Blank	0.000	06:33
Control Test	0.079	06:34
Air Blank	0.000	06:34
Diagnostics Check	OK	06:34

---

Data Download Date/Time: 11/07/2016 20:33

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: CRAMER, B  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 10/05/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 12716080A7  
DRY GAS EXPIRY DATE: 06/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 04:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:52

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 10/07/2016 06:20:01

---

Diagnostics Check	OK	06:12
Air Blank	0.000	06:13
Control Test	0.080	06:13
Air Blank	0.000	06:13
Subject Sample # 1	0.166	06:15
Breath Volume	1.324	06:15
Air Blank	0.000	06:16
Air Blank	0.000	06:17
Subject Sample # 2	0.183	06:18
Breath Volume	1.195	06:18
Air Blank	0.000	06:19
Control Test	0.078	06:19
Air Blank	0.000	06:19
Diagnostics Check	OK	06:20

---

Data Download Date/Time: 11/07/2016 20:33

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TORRES,JENIFER,  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: LARA,A  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 10/05/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 12716080A7  
DRY GAS EXPIRY DATE: 06/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 01:52  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:46

SUBJECT'S LAST NAME: DEAN

DATE/TIME OF TEST: 10/08/2016 03:13:49

---

Diagnostics Check	OK	03:07
Air Blank	0.000	03:07
Control Test	0.079	03:07
Air Blank	0.000	03:08
Subject Sample # 1	0.121	03:08
Breath Volume	1.820	03:08
Air Blank	0.000	03:09
Air Blank	0.000	03:11
Subject Sample # 2	0.130	03:12
Breath Volume	2.203	03:12
Air Blank	0.000	03:12
Control Test	0.079	03:13
Air Blank	0.000	03:13
Diagnostics Check	OK	03:13

---

Data Download Date/Time: 11/07/2016 20:33

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: WOODEN, D  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 10/05/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 12716080A7  
DRY GAS EXPIRY DATE: 06/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 04:22  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:48

SUBJECT'S LAST NAME: VELEZ DE GRACIA

DATE/TIME OF TEST: 10/08/2016 05:15:10

---

Diagnostics Check		OK	05:08
Air Blank		0.000	05:09
Control Test		0.079	05:09
Air Blank		0.000	05:10
Subject Sample	# 1	0.120	05:10
Breath Volume		3.417	05:10
Air Blank		0.000	05:11
Air Blank		0.000	05:13
Subject Sample	# 2	0.127	05:13
Breath Volume		3.921	05:13
Air Blank		0.000	05:14
Control Test		0.079	05:14
Air Blank		0.000	05:15
Diagnostics Check		OK	05:15

---

Data Download Date/Time: 11/07/2016 20:33



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MCMINN,KATHY,  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: PECYLAK,M  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 10/05/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 12716080A7  
DRY GAS EXPIRY DATE: 06/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 01:18  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: MONTOTO

DATE/TIME OF TEST: 10/09/2016 02:26:08

---

Diagnostics Check	OK	02:18
Air Blank	0.000	02:18
Control Test	0.080	02:19
Air Blank	0.000	02:19
Subject Sample # 1	0.000	02:21
Breath Volume	1.761	02:21
Air Blank	0.000	02:21
Air Blank	0.000	02:23
Subject Sample # 2	0.000	02:24
Breath Volume	1.402	02:24
Air Blank	0.000	02:25
Control Test	0.079	02:25
Air Blank	0.000	02:26
Diagnostics Check	OK	02:26

---

Data Download Date/Time: 11/07/2016 20:33

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KIRK,CASEY,S  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: SNAGGS,S  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 10/05/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 12716080A7  
DRY GAS EXPIRY DATE: 06/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 02:57  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:26

SUBJECT'S LAST NAME: HUMENIK

DATE/TIME OF TEST: 10/09/2016 03:53:00

---

Diagnostics Check		OK	03:46
Air Blank		0.000	03:47
Control Test		0.079	03:47
Air Blank		0.000	03:47
Subject Sample	# 1	0.284	03:48
Breath Volume		2.421	03:48
Air Blank		0.000	03:49
Air Blank		0.000	03:50
Subject Sample	# 2	0.285	03:51
Breath Volume		2.226	03:51
Air Blank		0.000	03:52
Control Test		0.080	03:52
Air Blank		0.000	03:52
Diagnostics Check		OK	03:53

---

Data Download Date/Time: 11/07/2016 20:33

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DAIL,JULIE,A  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: GONZALEZ,J  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 10/05/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 12716080A7  
DRY GAS EXPIRY DATE: 06/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 04:25  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:53

SUBJECT'S LAST NAME: CHASTEEN

DATE/TIME OF TEST: 10/09/2016 05:20:56

---

Diagnostics Check	OK	05:14
Air Blank	0.000	05:14
Control Test	0.079	05:14
Air Blank	0.000	05:15
Subject Sample # 1	0.156	05:16
Breath Volume	1.660	05:16
Air Blank	0.000	05:17
Air Blank	0.000	05:18
Subject Sample # 2	0.153	05:19
Breath Volume	1.542	05:19
Air Blank	0.000	05:20
Control Test	0.078	05:20
Air Blank	0.000	05:20
Diagnostics Check	OK	05:20

---

Data Download Date/Time: 11/07/2016 20:33

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: PECYLAK,M  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 10/05/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 12716080A7  
DRY GAS EXPIRY DATE: 06/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 22:35  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:13

SUBJECT'S LAST NAME: MOOERS

DATE/TIME OF TEST: 10/09/2016 23:41:49

---

Diagnostics Check	OK	23:33	
Air Blank	0.000	23:33	
Control Test	0.080	23:34	
Air Blank	0.000	23:34	
Subject Sample	# 1	23:37	No Sample Provided
Breath Volume	0.000	23:37	
Air Blank	0.000	23:38	
Air Blank	0.000	23:40	
Subject Sample	# 2	23:40	Subject Test Refused
Breath Volume	0.000	23:40	
Air Blank	0.000	23:40	
Control Test	0.079	23:41	
Air Blank	0.000	23:41	
Diagnostics Check	OK	23:41	

---

Data Download Date/Time: 11/07/2016 20:33

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MCMINN,KATHY,  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: LYON,M  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 10/05/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 12716080A7  
DRY GAS EXPIRY DATE: 06/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 03:32  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:09

SUBJECT'S LAST NAME: HINDS

DATE/TIME OF TEST: 10/10/2016 04:38:28

---

Diagnostics Check	OK	04:30
Air Blank	0.000	04:31
Control Test	0.080	04:31
Air Blank	0.000	04:32
Subject Sample # 1	0.198	04:33
Breath Volume	1.367	04:33
Air Blank	0.000	04:34
Air Blank	0.000	04:36
Subject Sample # 2	0.212	04:36
Breath Volume	1.136	04:36
Air Blank	0.000	04:37
Control Test	0.079	04:37
Air Blank	0.000	04:38
Diagnostics Check	OK	04:38

---

Data Download Date/Time: 11/07/2016 20:33

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DAIL,JULIE,A  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: FRANKLIN,R  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 10/05/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 12716080A7  
DRY GAS EXPIRY DATE: 06/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 01:41  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:07

SUBJECT'S LAST NAME: SOHL

DATE/TIME OF TEST: 10/11/2016 02:33:49

---

Diagnostics Check	OK	02:27
Air Blank	0.000	02:27
Control Test	0.079	02:28
Air Blank	0.000	02:28
Subject Sample # 1	0.186	02:29
Breath Volume	1.464	02:29
Air Blank	0.000	02:29
Air Blank	0.000	02:31
Subject Sample # 2	0.191	02:32
Breath Volume	1.890	02:32
Air Blank	0.000	02:32
Control Test	0.080	02:33
Air Blank	0.000	02:33
Diagnostics Check	OK	02:33

---

Data Download Date/Time: 11/07/2016 20:33

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KIRK,CASEY,S  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: DEJESUS,V  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 10/05/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 12716080A7  
DRY GAS EXPIRY DATE: 06/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 02:07  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:49

SUBJECT'S LAST NAME: TORRES

DATE/TIME OF TEST: 10/11/2016 03:16:01

---

Diagnostics Check	OK	03:09
Air Blank	0.000	03:10
Control Test	0.079	03:10
Air Blank	0.000	03:10
Subject Sample # 1	0.183	03:11
Breath Volume	2.343	03:11
Air Blank	0.000	03:12
Air Blank	0.000	03:13
Subject Sample # 2	0.202	03:14
Breath Volume	2.300	03:14
Air Blank	0.000	03:15
Control Test	0.079	03:15
Air Blank	0.000	03:15
Diagnostics Check	OK	03:16

---

Data Download Date/Time: 11/07/2016 20:33

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KIRK,CASEY,S  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: PORTMAN,A  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 10/05/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 12716080A7  
DRY GAS EXPIRY DATE: 06/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 00:21  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:14

SUBJECT'S LAST NAME: PENDLETON

DATE/TIME OF TEST: 10/14/2016 01:41:16

---

Diagnostics Check	OK	01:34
Air Blank	0.000	01:35
Control Test	0.080	01:35
Air Blank	0.000	01:36
Subject Sample # 1	0.000	01:36
Breath Volume	1.875	01:36
Air Blank	0.000	01:37
Air Blank	0.000	01:39
Subject Sample # 2	0.000	01:39
Breath Volume	2.542	01:39
Air Blank	0.000	01:40
Control Test	0.080	01:40
Air Blank	0.000	01:41
Diagnostics Check	OK	01:41

---

Data Download Date/Time: 11/07/2016 20:33



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: PECYLAK,M  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 10/05/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 12716080A7  
DRY GAS EXPIRY DATE: 06/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 02:42  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:22

SUBJECT'S LAST NAME: BRIMMER

DATE/TIME OF TEST: 10/14/2016 03:49:31

---

Diagnostics Check	OK	03:42
Air Blank	0.000	03:42
Control Test	0.080	03:43
Air Blank	0.000	03:43
Subject Sample # 1	0.187	03:44
Breath Volume	1.242	03:44
Air Blank	0.000	03:45
Air Blank	0.000	03:46
Subject Sample # 2	0.179	03:47
Breath Volume	1.824	03:47
Air Blank	0.000	03:48
Control Test	0.078	03:48
Air Blank	0.000	03:49
Diagnostics Check	OK	03:49

---

Data Download Date/Time: 11/07/2016 20:33

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KIRK,CASEY,S  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: SNAGGS,S  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 10/05/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 12716080A7  
DRY GAS EXPIRY DATE: 06/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 22:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:47

SUBJECT'S LAST NAME: HICKS

DATE/TIME OF TEST: 10/14/2016 23:14:49

---

Diagnostics Check	OK	23:08
Air Blank	0.000	23:09
Control Test	0.080	23:09
Air Blank	0.000	23:10
Subject Sample # 1	0.000	23:10
Breath Volume	3.988	23:10
Air Blank	0.000	23:10
Air Blank	0.000	23:12
Subject Sample # 2	0.000	23:13
Breath Volume	3.882	23:13
Air Blank	0.000	23:13
Control Test	0.079	23:14
Air Blank	0.000	23:14
Diagnostics Check	OK	23:14

---

Data Download Date/Time: 11/07/2016 20:33

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DAIL,JULIE,A  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: DAWSY,B  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 10/05/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 12716080A7  
DRY GAS EXPIRY DATE: 06/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 00:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:02

SUBJECT'S LAST NAME: HIGBEE

DATE/TIME OF TEST: 10/15/2016 01:32:06

---

Diagnostics Check	OK	01:25
Air Blank	0.000	01:26
Control Test	0.080	01:26
Air Blank	0.000	01:26
Subject Sample # 1	0.161	01:27
Breath Volume	2.050	01:27
Air Blank	0.000	01:28
Air Blank	0.000	01:30
Subject Sample # 2	0.168	01:30
Breath Volume	2.390	01:30
Air Blank	0.000	01:31
Control Test	0.079	01:31
Air Blank	0.000	01:32
Diagnostics Check	OK	01:32

---

Data Download Date/Time: 11/07/2016 20:33

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: FRANKLIN,R  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 10/05/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 12716080A7  
DRY GAS EXPIRY DATE: 06/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 02:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:29

SUBJECT'S LAST NAME: PERDOMO-SUAZO

DATE/TIME OF TEST: 10/15/2016 03:08:42

---

Diagnostics Check	OK	03:01
Air Blank	0.000	03:02
Control Test	0.080	03:02
Air Blank	0.000	03:02
Subject Sample # 1	0.194	03:04
Breath Volume	3.605	03:04
Air Blank	0.000	03:04
Air Blank	0.000	03:06
Subject Sample # 2	0.184	03:07
Breath Volume	2.121	03:07
Air Blank	0.000	03:07
Control Test	0.078	03:08
Air Blank	0.000	03:08
Diagnostics Check	OK	03:08

---

Data Download Date/Time: 11/07/2016 20:33

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DAIL,JULIE,A  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: GONZALEZ,D  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 10/05/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 12716080A7  
DRY GAS EXPIRY DATE: 06/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 02:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:26

SUBJECT'S LAST NAME: ELLIS

DATE/TIME OF TEST: 10/15/2016 03:53:33

---

Diagnostics Check	OK	03:47
Air Blank	0.000	03:47
Control Test	0.079	03:47
Air Blank	0.000	03:48
Subject Sample # 1	0.083	03:49
Breath Volume	1.179	03:49
Air Blank	0.000	03:49
Air Blank	0.000	03:51
Subject Sample # 2	0.079	03:51
Breath Volume	1.289	03:51
Air Blank	0.000	03:52
Control Test	0.079	03:52
Air Blank	0.000	03:53
Diagnostics Check	OK	03:53

---

Data Download Date/Time: 11/07/2016 20:33

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TORRES,JENIFER,  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: SNAGGS,S  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 10/05/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 12716080A7  
DRY GAS EXPIRY DATE: 06/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 03:12  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: FAUST

DATE/TIME OF TEST: 10/16/2016 04:11:53

---

Diagnostics Check	OK	04:05
Air Blank	0.000	04:06
Control Test	0.080	04:06
Air Blank	0.000	04:06
Subject Sample # 1	0.174	04:07
Breath Volume	1.906	04:07
Air Blank	0.000	04:08
Air Blank	0.000	04:09
Subject Sample # 2	0.189	04:10
Breath Volume	1.324	04:10
Air Blank	0.000	04:10
Control Test	0.079	04:11
Air Blank	0.000	04:11
Diagnostics Check	OK	04:11

---

Data Download Date/Time: 11/07/2016 20:33

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SIMMONS,ANGELA,V  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BARFIELD CRAIG,C  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 10/05/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 12716080A7  
DRY GAS EXPIRY DATE: 06/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 15:34  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 16:20

SUBJECT'S LAST NAME: MC DONALD

DATE/TIME OF TEST: 10/17/2016 16:52:24

---

Diagnostics Check	OK	16:40
Air Blank	0.000	16:41
Control Test	0.080	16:41
Air Blank	0.000	16:42
Subject Sample # 1	0.319	16:42
Breath Volume	1.589	16:42
Air Blank	0.000	16:43
Air Blank	0.000	16:45
Subject Sample # 2	0.291	16:45
Breath Volume	1.406	16:45
Air Blank	0.000	16:46
Air Blank	0.000	16:48
Subject Sample # 3	0.328	16:50
Breath Volume	1.632	16:50
Air Blank	0.000	16:51
Control Test	0.078	16:51
Air Blank	0.000	16:52
Diagnostics Check	OK	16:52

---

Data Download Date/Time: 11/07/2016 20:33

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BAILEY,D  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 10/05/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 12716080A7  
DRY GAS EXPIRY DATE: 06/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 01:44  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:04

SUBJECT'S LAST NAME: FIGUEROA

DATE/TIME OF TEST: 10/19/2016 02:30:36

---

Diagnostics Check	OK	02:24
Air Blank	0.000	02:24
Control Test	0.079	02:25
Air Blank	0.000	02:25
Subject Sample # 1	0.190	02:26
Breath Volume	2.937	02:26
Air Blank	0.000	02:26
Air Blank	0.000	02:28
Subject Sample # 2	0.196	02:28
Breath Volume	2.917	02:28
Air Blank	0.000	02:29
Control Test	0.079	02:29
Air Blank	0.000	02:30
Diagnostics Check	OK	02:30

---

Data Download Date/Time: 11/07/2016 20:33



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KIRK,CASEY,S  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: VAN-TREESE,S  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 10/05/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 12716080A7  
DRY GAS EXPIRY DATE: 06/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 01:46  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:17

SUBJECT'S LAST NAME: LA BONTE

DATE/TIME OF TEST: 10/20/2016 02:44:37

---

Diagnostics Check	OK	02:37
Air Blank	0.000	02:37
Control Test	0.079	02:38
Air Blank	0.000	02:38
Subject Sample # 1	0.008	02:40
Breath Volume	1.246	02:40
Air Blank	0.000	02:40
Air Blank	0.000	02:42
Subject Sample # 2	0.000	02:43
Breath Volume	1.144	02:43
Air Blank	0.000	02:43
Control Test	0.079	02:44
Air Blank	0.000	02:44
Diagnostics Check	OK	02:44

---

Data Download Date/Time: 11/07/2016 20:33

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DAIL,JULIE,A  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BARLAUG,G  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 10/05/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 12716080A7  
DRY GAS EXPIRY DATE: 06/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 02:10  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:47

SUBJECT'S LAST NAME: SIFFRINN

DATE/TIME OF TEST: 10/20/2016 03:15:50

---

Diagnostics Check	OK	03:07
Air Blank	0.000	03:08
Control Test	0.079	03:08
Air Blank	0.000	03:09
Subject Sample # 1	0.138	03:10
Breath Volume	1.644	03:10
Air Blank	0.000	03:11
Air Blank	0.000	03:13
Subject Sample # 2	0.138	03:14
Breath Volume	1.234	03:14
Air Blank	0.000	03:14
Control Test	0.078	03:15
Air Blank	0.000	03:15
Diagnostics Check	OK	03:15

---

Data Download Date/Time: 11/07/2016 20:33

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BARLAUG, G  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 10/05/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 12716080A7  
DRY GAS EXPIRY DATE: 06/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 02:51  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:13

SUBJECT'S LAST NAME: GORDON

DATE/TIME OF TEST: 10/21/2016 03:40:47

---

Diagnostics Check	OK	03:34
Air Blank	0.000	03:34
Control Test	0.079	03:34
Air Blank	0.000	03:35
Subject Sample # 1	0.151	03:35
Breath Volume	1.878	03:35
Air Blank	0.000	03:36
Air Blank	0.000	03:38
Subject Sample # 2	0.155	03:39
Breath Volume	2.730	03:39
Air Blank	0.000	03:39
Control Test	0.079	03:40
Air Blank	0.000	03:40
Diagnostics Check	OK	03:40

---

Data Download Date/Time: 11/07/2016 20:33

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HENRY,PATRICIA,  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: GONZALEZ,J  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 10/05/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 12716080A7  
DRY GAS EXPIRY DATE: 06/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 22:56  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:06

SUBJECT'S LAST NAME: GRIMES

DATE/TIME OF TEST: 10/21/2016 23:32:48

---

Diagnostics Check	OK	23:26
Air Blank	0.000	23:26
Control Test	0.080	23:26
Air Blank	0.000	23:27
Subject Sample # 1	0.179	23:28
Breath Volume	2.578	23:28
Air Blank	0.000	23:28
Air Blank	0.000	23:30
Subject Sample # 2	0.186	23:31
Breath Volume	3.609	23:31
Air Blank	0.000	23:31
Control Test	0.079	23:32
Air Blank	0.000	23:32
Diagnostics Check	OK	23:32

---

Data Download Date/Time: 11/07/2016 20:33

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: SIMPKINS,D  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 10/05/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 12716080A7  
DRY GAS EXPIRY DATE: 06/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 01:07  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:48

SUBJECT'S LAST NAME: FAN

DATE/TIME OF TEST: 10/23/2016 02:14:55

---

Diagnostics Check	OK	02:08
Air Blank	0.000	02:08
Control Test	0.080	02:09
Air Blank	0.000	02:09
Subject Sample # 1	0.145	02:10
Breath Volume	1.535	02:10
Air Blank	0.000	02:11
Air Blank	0.000	02:12
Subject Sample # 2	0.147	02:13
Breath Volume	1.636	02:13
Air Blank	0.000	02:13
Control Test	0.079	02:14
Air Blank	0.000	02:14
Diagnostics Check	OK	02:14

---

Data Download Date/Time: 11/07/2016 20:33

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DAIL,JULIE,A  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BAILEY,D  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 10/05/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 12716080A7  
DRY GAS EXPIRY DATE: 06/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 03:40  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: ROMERO-PENSO

DATE/TIME OF TEST: 10/23/2016 04:26:53

---

Diagnostics Check	OK	04:20
Air Blank	0.000	04:20
Control Test	0.080	04:21
Air Blank	0.000	04:21
Subject Sample # 1	0.176	04:22
Breath Volume	1.417	04:22
Air Blank	0.000	04:23
Air Blank	0.000	04:24
Subject Sample # 2	0.170	04:25
Breath Volume	1.882	04:25
Air Blank	0.000	04:25
Control Test	0.079	04:26
Air Blank	0.000	04:26
Diagnostics Check	OK	04:26

---

Data Download Date/Time: 11/07/2016 20:33

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KIRK,CASEY,S  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: WOODEN,D  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 10/05/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 12716080A7  
DRY GAS EXPIRY DATE: 06/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 23:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:28

SUBJECT'S LAST NAME: CACERES JR

DATE/TIME OF TEST: 10/23/2016 23:54:57

---

Diagnostics Check	OK	23:48
Air Blank	0.000	23:48
Control Test	0.080	23:49
Air Blank	0.000	23:49
Subject Sample # 1	0.038	23:50
Breath Volume	1.496	23:50
Air Blank	0.000	23:50
Air Blank	0.000	23:52
Subject Sample # 2	0.042	23:53
Breath Volume	3.792	23:53
Air Blank	0.000	23:54
Control Test	0.080	23:54
Air Blank	0.000	23:54
Diagnostics Check	OK	23:54

---

Data Download Date/Time: 11/07/2016 20:33

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TORRES,JENIFER,  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BARLAUG,G  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 10/05/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 12716080A7  
DRY GAS EXPIRY DATE: 06/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 03:47  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: MECONNAHEY

DATE/TIME OF TEST: 10/27/2016 04:37:01

---

Diagnostics Check	OK	04:30
Air Blank	0.000	04:30
Control Test	0.080	04:31
Air Blank	0.000	04:31
Subject Sample # 1	0.170	04:32
Breath Volume	2.683	04:32
Air Blank	0.000	04:32
Air Blank	0.000	04:34
Subject Sample # 2	0.172	04:35
Breath Volume	3.085	04:35
Air Blank	0.000	04:36
Control Test	0.079	04:36
Air Blank	0.000	04:36
Diagnostics Check	OK	04:37

---

Data Download Date/Time: 11/07/2016 20:33



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MCMINN,KATHY,  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BRENNAN,D  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 10/05/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 12716080A7  
DRY GAS EXPIRY DATE: 06/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 02:31  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: MONTERO

DATE/TIME OF TEST: 10/29/2016 03:17:45

---

Diagnostics Check	OK	03:11
Air Blank	0.000	03:11
Control Test	0.079	03:11
Air Blank	0.000	03:12
Subject Sample # 1	0.144	03:13
Breath Volume	1.441	03:13
Air Blank	0.000	03:13
Air Blank	0.000	03:15
Subject Sample # 2	0.152	03:16
Breath Volume	3.335	03:16
Air Blank	0.000	03:16
Control Test	0.078	03:17
Air Blank	0.000	03:17
Diagnostics Check	OK	03:17

---

Data Download Date/Time: 11/07/2016 20:33

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TORRES,JENIFER,  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: DAWSY,B  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 10/05/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 12716080A7  
DRY GAS EXPIRY DATE: 06/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 02:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:36

SUBJECT'S LAST NAME: CARRAZANA RAMIREZ

DATE/TIME OF TEST: 10/30/2016 03:03:59

---

Diagnostics Check	OK	02:57
Air Blank	0.000	02:57
Control Test	0.079	02:57
Air Blank	0.000	02:58
Subject Sample # 1	0.100	02:59
Breath Volume	2.281	02:59
Air Blank	0.000	02:59
Air Blank	0.000	03:01
Subject Sample # 2	0.092	03:02
Breath Volume	1.582	03:02
Air Blank	0.000	03:03
Control Test	0.079	03:03
Air Blank	0.000	03:03
Diagnostics Check	OK	03:03

---

Data Download Date/Time: 11/07/2016 20:33

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MEEHAN,SHANE,P  
BREATH TEST OPERATOR AGENCY: Flagler County SO  
ARRESTING OFFICER: TOWNS,R  
ARRESTING OFFICER AGENCY: Flagler County SO  
LAST AGENCY INSPECTION: 09/30/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28515080A3  
DRY GAS EXPIRY DATE: 11/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 06:46  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 07:59

SUBJECT'S LAST NAME: VERRECHIO

DATE/TIME OF TEST: 10/01/2016 08:26:06

---

Diagnostics Check	OK	08:19
Air Blank	0.000	08:20
Control Test	0.078	08:20
Air Blank	0.000	08:21
Subject Sample # 1	0.174	08:21
Breath Volume	2.636	08:21
Air Blank	0.000	08:22
Air Blank	0.000	08:24
Subject Sample # 2	0.175	08:24
Breath Volume	2.351	08:24
Air Blank	0.000	08:25
Control Test	0.078	08:25
Air Blank	0.000	08:26
Diagnostics Check	OK	08:26

---

Data Download Date/Time: 10/28/2016 13:33

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BEAUSOLEIL,AARON,F  
BREATH TEST OPERATOR AGENCY: Flagler County SO  
ARRESTING OFFICER: DUENAS,JONATHON  
ARRESTING OFFICER AGENCY: Flagler County SO  
LAST AGENCY INSPECTION: 09/30/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28515080A3  
DRY GAS EXPIRY DATE: 11/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 20:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:03

SUBJECT'S LAST NAME: ROBERTS

DATE/TIME OF TEST: 10/01/2016 21:30:27

---

Diagnostics Check	OK	21:23
Air Blank	0.000	21:24
Control Test	0.079	21:24
Air Blank	0.000	21:25
Subject Sample # 1	0.100	21:25
Breath Volume	2.105	21:25
Air Blank	0.000	21:26
Air Blank	0.000	21:28
Subject Sample # 2	0.095	21:28
Breath Volume	1.667	21:28
Air Blank	0.000	21:29
Control Test	0.079	21:29
Air Blank	0.000	21:30
Diagnostics Check	OK	21:30

---

Data Download Date/Time: 10/28/2016 13:33

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BEAUSOLEIL,AARON,F  
BREATH TEST OPERATOR AGENCY: Flagler County SO  
ARRESTING OFFICER: STOGDON,BRAD  
ARRESTING OFFICER AGENCY: Flagler County SO  
LAST AGENCY INSPECTION: 09/30/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28515080A3  
DRY GAS EXPIRY DATE: 11/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 00:43  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:59

SUBJECT'S LAST NAME: KARDAS

DATE/TIME OF TEST: 10/02/2016 02:23:32

---

Diagnostics Check	OK	02:20	
Air Blank	0.000	02:20	
Control Test	0.078	02:21	
Air Blank	0.000	02:21	
Subject Sample	# 1	02:22	Subject Test Refused
Breath Volume	0.000	02:22	
Air Blank	0.000	02:22	
Control Test	0.078	02:22	
Air Blank	0.000	02:23	
Diagnostics Check	OK	02:23	

---

Data Download Date/Time: 10/28/2016 13:33

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PRIEMER,AMANDA,M  
BREATH TEST OPERATOR AGENCY: Flagler County SO  
ARRESTING OFFICER: DUENAS,JONATHON  
ARRESTING OFFICER AGENCY: Flagler County SO  
LAST AGENCY INSPECTION: 09/30/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28515080A3  
DRY GAS EXPIRY DATE: 11/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 20:37  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:59

SUBJECT'S LAST NAME: MARTIN

DATE/TIME OF TEST: 10/02/2016 21:24:32

---

Diagnostics Check	OK	21:21	
Air Blank	0.000	21:22	
Control Test	0.079	21:22	
Air Blank	0.000	21:22	
Subject Sample	# 1	21:23	Subject Test Refused
Breath Volume	0.000	21:23	
Air Blank	0.000	21:23	
Control Test	0.078	21:23	
Air Blank	0.000	21:24	
Diagnostics Check	OK	21:24	

---

Data Download Date/Time: 10/28/2016 13:33

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FAIRCLOTH, EDWARD, W  
BREATH TEST OPERATOR AGENCY: Flagler County SO  
ARRESTING OFFICER: MEEHAN, SHANE  
ARRESTING OFFICER AGENCY: Flagler County SO  
LAST AGENCY INSPECTION: 09/30/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28515080A3  
DRY GAS EXPIRY DATE: 11/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 16:03  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:45

SUBJECT'S LAST NAME: FIELDS

DATE/TIME OF TEST: 10/05/2016 17:11:58

---

Diagnostics Check	OK	17:05
Air Blank	0.000	17:06
Control Test	0.079	17:06
Air Blank	0.000	17:06
Subject Sample # 1	0.055	17:07
Breath Volume	1.562	17:07
Air Blank	0.000	17:07
Air Blank	0.000	17:09
Subject Sample # 2	0.055	17:10
Breath Volume	1.496	17:10
Air Blank	0.000	17:11
Control Test	0.078	17:11
Air Blank	0.000	17:11
Diagnostics Check	OK	17:11

---

Data Download Date/Time: 10/28/2016 13:33

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FAIRCLOTH, EDWARD, W  
BREATH TEST OPERATOR AGENCY: Flagler County SO  
ARRESTING OFFICER: VINCL, R  
ARRESTING OFFICER AGENCY: Flagler County SO  
LAST AGENCY INSPECTION: 09/30/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28515080A3  
DRY GAS EXPIRY DATE: 11/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 13:40  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 14:25

SUBJECT'S LAST NAME: PROVAN

DATE/TIME OF TEST: 10/06/2016 14:51:50

---

Diagnostics Check	OK	14:45
Air Blank	0.000	14:45
Control Test	0.078	14:45
Air Blank	0.000	14:46
Subject Sample # 1	0.257	14:47
Breath Volume	2.125	14:47
Air Blank	0.000	14:47
Air Blank	0.000	14:49
Subject Sample # 2	0.254	14:50
Breath Volume	2.054	14:50
Air Blank	0.000	14:50
Control Test	0.078	14:51
Air Blank	0.000	14:51
Diagnostics Check	OK	14:51

---

Data Download Date/Time: 10/28/2016 13:33



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PRIEMER,AMANDA,M  
BREATH TEST OPERATOR AGENCY: Flagler County SO  
ARRESTING OFFICER: STOGDON,BRAD  
ARRESTING OFFICER AGENCY: Flagler County SO  
LAST AGENCY INSPECTION: 09/30/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28515080A3  
DRY GAS EXPIRY DATE: 11/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 22:19  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:13

SUBJECT'S LAST NAME: EDWARD

DATE/TIME OF TEST: 10/07/2016 23:41:09

---

Diagnostics Check	OK	23:35	
Air Blank	0.000	23:35	
Control Test	0.078	23:36	
Air Blank	0.000	23:36	
Subject Sample # 1	0.000	23:37	
Breath Volume	1.792	23:37	
Air Blank	0.000	23:37	
Air Blank	0.000	23:39	
Subject Sample # 2		23:39	RFI Detect
Breath Volume	0.000	23:39	
Air Blank	0.000	23:40	
Control Test	0.077	23:40	
Air Blank	0.000	23:41	
Diagnostics Check	OK	23:41	

---

Data Download Date/Time: 10/28/2016 13:33

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PRIEMER,AMANDA,M  
BREATH TEST OPERATOR AGENCY: Flagler County SO  
ARRESTING OFFICER: STOGDON,BRAD  
ARRESTING OFFICER AGENCY: Flagler County SO  
LAST AGENCY INSPECTION: 09/30/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28515080A3  
DRY GAS EXPIRY DATE: 11/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 22:19  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:13

SUBJECT'S LAST NAME: EDWARD

DATE/TIME OF TEST: 10/08/2016 00:02:51

---

Diagnostics Check	OK	23:55
Air Blank	0.000	23:56
Control Test	0.078	23:56
Air Blank	0.000	23:57
Subject Sample # 2	0.000	23:57
Breath Volume	2.179	23:57
Air Blank	0.000	23:58
Air Blank	0.000	00:00
Subject Sample # 1	0.000	00:01
Breath Volume	1.847	00:01
Air Blank	0.000	00:01
Control Test	0.077	00:02
Air Blank	0.000	00:02
Diagnostics Check	OK	00:02

---

Data Download Date/Time: 10/28/2016 13:33

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PRIEMER,AMANDA,M  
BREATH TEST OPERATOR AGENCY: Flagler County SO  
ARRESTING OFFICER: PETKOVSEK,RICHARD  
ARRESTING OFFICER AGENCY: Flagler County SO  
LAST AGENCY INSPECTION: 09/30/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28515080A3  
DRY GAS EXPIRY DATE: 11/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 03:32  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:59

SUBJECT'S LAST NAME: SIMMONS

DATE/TIME OF TEST: 10/08/2016 04:27:14

---

Diagnostics Check	OK	04:24	
Air Blank	0.000	04:24	
Control Test	0.078	04:24	
Air Blank	0.000	04:25	
Subject Sample	# 1	04:25	Subject Test Refused
Breath Volume	0.000	04:25	
Air Blank	0.000	04:26	
Control Test	0.078	04:26	
Air Blank	0.000	04:27	
Diagnostics Check	OK	04:27	

---

Data Download Date/Time: 10/28/2016 13:33

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PARKER,CARL,L  
BREATH TEST OPERATOR AGENCY: Flagler County SO  
ARRESTING OFFICER: PARKER,CARL  
ARRESTING OFFICER AGENCY: Flagler County SO  
LAST AGENCY INSPECTION: 09/30/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28515080A3  
DRY GAS EXPIRY DATE: 11/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 04:47  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:50

SUBJECT'S LAST NAME: LAROU

DATE/TIME OF TEST: 10/09/2016 05:36:33

---

Diagnostics Check	OK	05:33	
Air Blank	0.000	05:34	
Control Test	0.075	05:34	
Air Blank	0.000	05:34	
Subject Sample	# 1	05:35	Subject Test Refused
Breath Volume	0.000	05:35	
Air Blank	0.000	05:35	
Control Test	0.076	05:35	
Air Blank	0.000	05:36	
Diagnostics Check	OK	05:36	

---

Data Download Date/Time: 10/28/2016 13:33

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PARKER,CARL,L  
BREATH TEST OPERATOR AGENCY: Flagler County SO  
ARRESTING OFFICER: PETKOVSEK,RICHARD  
ARRESTING OFFICER AGENCY: Flagler County SO  
LAST AGENCY INSPECTION: 09/30/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28515080A3  
DRY GAS EXPIRY DATE: 11/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 23:12  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:38

SUBJECT'S LAST NAME: BUJNOWSKI

DATE/TIME OF TEST: 10/10/2016 00:05:06

---

Diagnostics Check	OK	00:02	
Air Blank	0.000	00:02	
Control Test	0.079	00:03	
Air Blank	0.000	00:03	
Subject Sample	# 1	00:03	Subject Test Refused
Breath Volume	0.000	00:03	
Air Blank	0.000	00:04	
Control Test	0.079	00:04	
Air Blank	0.000	00:05	
Diagnostics Check	OK	00:05	

---

Data Download Date/Time: 10/28/2016 13:33

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R  
BREATH TEST OPERATOR AGENCY: FHP Troop G  
ARRESTING OFFICER: MONTGOMERY,KENNETH  
ARRESTING OFFICER AGENCY: FHP Troop G  
LAST AGENCY INSPECTION: 09/30/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28515080A3  
DRY GAS EXPIRY DATE: 11/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 02:28  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 10/11/2016 03:56:36

---

Diagnostics Check	OK	03:50
Air Blank	0.000	03:50
Control Test	0.079	03:51
Air Blank	0.000	03:51
Subject Sample # 1	0.231	03:52
Breath Volume	3.312	03:52
Air Blank	0.000	03:52
Air Blank	0.000	03:54
Subject Sample # 2	0.226	03:55
Breath Volume	3.277	03:55
Air Blank	0.000	03:55
Control Test	0.078	03:56
Air Blank	0.000	03:56
Diagnostics Check	OK	03:56

---

Data Download Date/Time: 10/28/2016 13:33

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DESCARTES,PETER,J  
BREATH TEST OPERATOR AGENCY: Flagler County SO  
ARRESTING OFFICER: PETKOVSEK,RICHARD  
ARRESTING OFFICER AGENCY: Flagler County SO  
LAST AGENCY INSPECTION: 09/30/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28515080A3  
DRY GAS EXPIRY DATE: 11/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 02:14  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:14

SUBJECT'S LAST NAME: SANTORA

DATE/TIME OF TEST: 10/12/2016 02:45:38

---

Diagnostics Check	OK	02:39
Air Blank	0.000	02:39
Control Test	0.079	02:40
Air Blank	0.000	02:40
Subject Sample # 1	0.000	02:41
Breath Volume	2.015	02:41
Air Blank	0.000	02:41
Air Blank	0.000	02:43
Subject Sample # 2	0.000	02:44
Breath Volume	1.546	02:44
Air Blank	0.000	02:44
Control Test	0.077	02:45
Air Blank	0.000	02:45
Diagnostics Check	OK	02:45

---

Data Download Date/Time: 10/28/2016 13:33

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILLIAMS,JASON,B  
BREATH TEST OPERATOR AGENCY: Flagler County SO  
ARRESTING OFFICER: COZZONE,G  
ARRESTING OFFICER AGENCY: Flagler Beach PD  
LAST AGENCY INSPECTION: 09/30/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28515080A3  
DRY GAS EXPIRY DATE: 11/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 22:12  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:01

SUBJECT'S LAST NAME: CZISNY

DATE/TIME OF TEST: 10/13/2016 23:33:01

---

Diagnostics Check	OK	23:21	
Air Blank	0.000	23:22	
Control Test	0.079	23:22	
Air Blank	0.000	23:22	
Subject Sample	# 1	23:26	No Sample Provided
Breath Volume	0.000	23:26	
Air Blank	0.000	23:26	
Air Blank	0.000	23:28	
Subject Sample	# 2	23:31	No Sample Provided
Breath Volume	0.000	23:31	
Air Blank	0.000	23:32	
Control Test	0.079	23:32	
Air Blank	0.000	23:32	
Diagnostics Check	OK	23:33	

---

Data Download Date/Time: 10/28/2016 13:33



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PRIEMER,AMANDA,M  
BREATH TEST OPERATOR AGENCY: Flagler County SO  
ARRESTING OFFICER: DUENAS,JONATHON  
ARRESTING OFFICER AGENCY: Flagler County SO  
LAST AGENCY INSPECTION: 09/30/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28515080A3  
DRY GAS EXPIRY DATE: 11/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 00:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:38

SUBJECT'S LAST NAME: HOWARD

DATE/TIME OF TEST: 10/17/2016 01:06:45

---

Diagnostics Check	OK	01:02	
Air Blank	0.000	01:02	
Control Test	0.078	01:03	
Air Blank	0.000	01:03	
Subject Sample	# 1	01:05	Slope Not Met
Breath Volume	1.046	01:05	
Air Blank	0.000	01:05	
Control Test	0.078	01:06	
Air Blank	0.000	01:06	
Diagnostics Check	OK	01:06	

---

Data Download Date/Time: 10/28/2016 13:33

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PRIEMER,AMANDA,M  
BREATH TEST OPERATOR AGENCY: Flagler County SO  
ARRESTING OFFICER: DUENAS,JONATHON  
ARRESTING OFFICER AGENCY: Flagler County SO  
LAST AGENCY INSPECTION: 09/30/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28515080A3  
DRY GAS EXPIRY DATE: 11/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 00:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:38

SUBJECT'S LAST NAME: HOWARD

DATE/TIME OF TEST: 10/17/2016 01:15:51

---

Diagnostics Check	OK	01:08
Air Blank	0.000	01:08
Control Test	0.078	01:09
Air Blank	0.000	01:09
Subject Sample # 1	0.197	01:10
Breath Volume	1.562	01:10
Air Blank	0.000	01:11
Air Blank	0.000	01:13
Subject Sample # 2	0.195	01:14
Breath Volume	2.031	01:14
Air Blank	0.000	01:14
Control Test	0.076	01:15
Air Blank	0.000	01:15
Diagnostics Check	OK	01:15

---

Data Download Date/Time: 10/28/2016 13:33

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PARKER,CARL,L  
BREATH TEST OPERATOR AGENCY: Flagler County SO  
ARRESTING OFFICER: PETKOVSEK,RICHARD  
ARRESTING OFFICER AGENCY: Flagler County SO  
LAST AGENCY INSPECTION: 09/30/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28515080A3  
DRY GAS EXPIRY DATE: 11/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 03:32  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: SALOM

DATE/TIME OF TEST: 10/23/2016 04:21:54

---

Diagnostics Check	OK	04:18	
Air Blank	0.000	04:19	
Control Test	0.079	04:19	
Air Blank	0.000	04:20	
Subject Sample	# 1	04:20	Subject Test Refused
Breath Volume	0.000	04:20	
Air Blank	0.000	04:20	
Control Test	0.079	04:21	
Air Blank	0.000	04:21	
Diagnostics Check	OK	04:21	

---

Data Download Date/Time: 10/28/2016 13:33

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PREVATT,JENNIFER,M  
BREATH TEST OPERATOR AGENCY: Flagler County SO  
ARRESTING OFFICER: SMITH,J  
ARRESTING OFFICER AGENCY: Flagler County SO  
LAST AGENCY INSPECTION: 09/30/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28515080A3  
DRY GAS EXPIRY DATE: 11/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 16:04  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 16:32

SUBJECT'S LAST NAME: FOX

DATE/TIME OF TEST: 10/23/2016 17:02:36

---

Diagnostics Check	OK	16:55
Air Blank	0.000	16:56
Control Test	0.080	16:56
Air Blank	0.000	16:57
Subject Sample # 1	0.000	16:58
Breath Volume	3.136	16:58
Air Blank	0.000	16:58
Air Blank	0.000	17:00
Subject Sample # 2	0.000	17:01
Breath Volume	3.632	17:01
Air Blank	0.000	17:01
Control Test	0.080	17:01
Air Blank	0.000	17:02
Diagnostics Check	OK	17:02

---

Data Download Date/Time: 10/28/2016 13:33

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PARKER,CARL,L  
BREATH TEST OPERATOR AGENCY: Flagler County SO  
ARRESTING OFFICER: PRIESTER,PAULA  
ARRESTING OFFICER AGENCY: Flagler County SO  
LAST AGENCY INSPECTION: 09/30/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28515080A3  
DRY GAS EXPIRY DATE: 11/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 03:20  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: MCDONALD

DATE/TIME OF TEST: 10/24/2016 04:08:02

---

Diagnostics Check	OK	04:02	
Air Blank	0.000	04:03	
Control Test	0.080	04:03	
Air Blank	0.000	04:04	
Subject Sample # 1	0.273	04:04	
Breath Volume	1.929	04:04	
Air Blank	0.000	04:05	
Air Blank		04:07	Ambient Fail
Air Blank	0.000	04:08	

---

Data Download Date/Time: 10/28/2016 13:33

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PARKER,CARL,L  
BREATH TEST OPERATOR AGENCY: Flagler County SO  
ARRESTING OFFICER: PRIESTER,PAULA  
ARRESTING OFFICER AGENCY: Flagler County SO  
LAST AGENCY INSPECTION: 09/30/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28515080A3  
DRY GAS EXPIRY DATE: 11/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 03:20  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: MCDONALD

DATE/TIME OF TEST: 10/24/2016 04:17:25

---

Diagnostics Check		OK	04:10
Air Blank		0.000	04:10
Control Test		0.078	04:11
Air Blank		0.000	04:11
Subject Sample	# 1	0.288	04:12
Breath Volume		2.179	04:12
Air Blank		0.000	04:13
Air Blank		0.000	04:15
Subject Sample	# 2	0.282	04:15
Breath Volume		1.257	04:15
Air Blank		0.000	04:16
Control Test		0.077	04:16
Air Blank		0.000	04:17
Diagnostics Check		OK	04:17

---

Data Download Date/Time: 10/28/2016 13:33

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PARKER,CARL,L  
BREATH TEST OPERATOR AGENCY: Flagler County SO  
ARRESTING OFFICER: PARKER,CARL  
ARRESTING OFFICER AGENCY: Flagler County SO  
LAST AGENCY INSPECTION: 09/30/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28515080A3  
DRY GAS EXPIRY DATE: 11/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 00:12  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: RHODES

DATE/TIME OF TEST: 10/27/2016 01:05:37

---

Diagnostics Check	OK	01:02	
Air Blank	0.000	01:03	
Control Test	0.079	01:03	
Air Blank	0.000	01:03	
Subject Sample	# 1	01:04	Subject Test Refused
Breath Volume	0.000	01:04	
Air Blank	0.000	01:04	
Control Test	0.078	01:04	
Air Blank	0.000	01:05	
Diagnostics Check	OK	01:05	

---

Data Download Date/Time: 10/28/2016 13:33

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PARKER,CARL,L  
BREATH TEST OPERATOR AGENCY: Flagler County SO  
ARRESTING OFFICER: PETKOVSEK,RICHARD  
ARRESTING OFFICER AGENCY: Flagler County SO  
LAST AGENCY INSPECTION: 09/30/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28515080A3  
DRY GAS EXPIRY DATE: 11/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 01:32  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:17

SUBJECT'S LAST NAME: WELLS

DATE/TIME OF TEST: 10/27/2016 02:43:26

---

Diagnostics Check	OK	02:39	
Air Blank	0.000	02:40	
Control Test	0.079	02:40	
Air Blank	0.000	02:41	
Subject Sample	# 1	02:42	Subject Test Refused
Breath Volume	0.000	02:42	
Air Blank	0.000	02:42	
Control Test	0.078	02:42	
Air Blank	0.000	02:43	
Diagnostics Check	OK	02:43	

---

Data Download Date/Time: 10/28/2016 13:33



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MEEHAN,SHANE,P  
BREATH TEST OPERATOR AGENCY: Flagler County SO  
ARRESTING OFFICER: MEEHAN,SHANE  
ARRESTING OFFICER AGENCY: Flagler County SO  
LAST AGENCY INSPECTION: 09/30/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28515080A3  
DRY GAS EXPIRY DATE: 11/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 08:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 09:38

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 10/28/2016 10:05:10

---

Diagnostics Check	OK	09:58
Air Blank	0.000	09:59
Control Test	0.079	09:59
Air Blank	0.000	10:00
Subject Sample # 1	0.000	10:00
Breath Volume	2.039	10:00
Air Blank	0.000	10:01
Air Blank	0.000	10:03
Subject Sample # 2	0.000	10:03
Breath Volume	1.886	10:03
Air Blank	0.000	10:04
Control Test	0.079	10:04
Air Blank	0.000	10:05
Diagnostics Check	OK	10:05

---

Data Download Date/Time: 10/28/2016 13:33

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BEAUSOLEIL,AARON,F  
BREATH TEST OPERATOR AGENCY: Flagler County SO  
ARRESTING OFFICER: PEDERSEN,ERIK  
ARRESTING OFFICER AGENCY: Flagler County SO  
LAST AGENCY INSPECTION: 10/28/2016

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 28515080A3  
DRY GAS EXPIRY DATE: 11/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 22:11  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:47

SUBJECT'S LAST NAME: MESERVE

DATE/TIME OF TEST: 10/29/2016 23:16:03

---

Diagnostics Check	OK	23:08	
Air Blank	0.000	23:09	
Control Test	0.079	23:09	
Air Blank	0.000	23:10	
Subject Sample # 1	0.198	23:11	
Breath Volume	2.332	23:11	
Air Blank	0.000	23:12	
Air Blank	0.000	23:14	
Subject Sample # 2		23:14	RFI Detect
Breath Volume	0.792	23:14	
Air Blank	0.000	23:15	
Control Test	0.078	23:15	
Air Blank	0.000	23:15	
Diagnostics Check	OK	23:16	

---

Data Download Date/Time: 11/22/2016 11:59

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006628

INSTRUMENT REGISTERED TO: FHP  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ZITO,ANTHONY,P  
BREATH TEST OPERATOR AGENCY: FHP Troop D  
ARRESTING OFFICER: ZITO,ANTHONY  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 09/30/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06616080A2  
DRY GAS EXPIRY DATE: 04/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 04:56  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:03

SUBJECT'S LAST NAME: PADILLA

DATE/TIME OF TEST: 10/21/2016 05:34:41

---

Diagnostics Check	OK	05:33	
Air Blank	0.000	05:33	
Control Test	0.000	05:34	Control Outside Tolerance
Air Blank	0.000	05:34	

---

Data Download Date/Time: 10/22/2016 21:36

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006628

INSTRUMENT REGISTERED TO: FHP  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ZITO,ANTHONY,P  
BREATH TEST OPERATOR AGENCY: FHP Troop D  
ARRESTING OFFICER: ZITO,ANTHONY  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 09/30/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06616080A2  
DRY GAS EXPIRY DATE: 04/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 04:56  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:03

SUBJECT'S LAST NAME: PADILLA

DATE/TIME OF TEST: 10/21/2016 05:44:56

---

Diagnostics Check	OK	05:36
Air Blank	0.000	05:36
Control Test	0.077	05:36
Air Blank	0.000	05:37
Subject Sample # 1	0.149	05:40
Breath Volume	1.695	05:40
Air Blank	0.000	05:41
Air Blank	0.000	05:42
Subject Sample # 2	0.150	05:42
Breath Volume	1.675	05:42
Air Blank	0.000	05:43
Control Test	0.075	05:44
Air Blank	0.000	05:44
Diagnostics Check	OK	05:44

---

Data Download Date/Time: 10/22/2016 21:36

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006629

INSTRUMENT REGISTERED TO: FHP  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HAWKINS,DEBORAH,  
BREATH TEST OPERATOR AGENCY: FHP Troop D  
ARRESTING OFFICER: HAWKINS,DEBORAH  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 09/30/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 09216080A3  
DRY GAS EXPIRY DATE: 05/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 21:55  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:26

SUBJECT'S LAST NAME: BELLINO

DATE/TIME OF TEST: 10/07/2016 22:58:04

---

Diagnostics Check	OK	22:51
Air Blank	0.000	22:51
Control Test	0.078	22:52
Air Blank	0.000	22:52
Subject Sample # 1	0.081	22:53
Breath Volume	2.738	22:53
Air Blank	0.000	22:53
Air Blank	0.000	22:55
Subject Sample # 2	0.077	22:56
Breath Volume	2.003	22:56
Air Blank	0.000	22:57
Control Test	0.078	22:57
Air Blank	0.000	22:57
Diagnostics Check	OK	22:58

---

Data Download Date/Time: 10/22/2016 21:31

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006629

INSTRUMENT REGISTERED TO: FHP  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HAWKINS,DEBORAH,  
BREATH TEST OPERATOR AGENCY: FHP Troop D  
ARRESTING OFFICER: HAWKINS,DEBORAH  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 09/30/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 09216080A3  
DRY GAS EXPIRY DATE: 05/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 03:08  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:22

SUBJECT'S LAST NAME: CUENCA

DATE/TIME OF TEST: 10/08/2016 03:52:08

---

Diagnostics Check	OK	03:45
Air Blank	0.000	03:46
Control Test	0.079	03:46
Air Blank	0.000	03:46
Subject Sample # 1	0.101	03:47
Breath Volume	2.179	03:47
Air Blank	0.000	03:48
Air Blank	0.000	03:50
Subject Sample # 2	0.096	03:50
Breath Volume	2.128	03:50
Air Blank	0.000	03:51
Control Test	0.077	03:51
Air Blank	0.000	03:52
Diagnostics Check	OK	03:52

---

Data Download Date/Time: 10/22/2016 21:31

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006629

INSTRUMENT REGISTERED TO: FHP  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HAWKINS,DEBORAH,  
BREATH TEST OPERATOR AGENCY: FHP Troop D  
ARRESTING OFFICER: ROBLEDO,MIGUEL  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 09/30/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 09216080A3  
DRY GAS EXPIRY DATE: 05/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 02:19  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:56

SUBJECT'S LAST NAME: DIAZ VELASQUEZ

DATE/TIME OF TEST: 10/09/2016 03:29:50

---

Diagnostics Check	OK	03:21
Air Blank	0.000	03:22
Control Test	0.078	03:22
Air Blank	0.000	03:23
Subject Sample # 1	0.176	03:25
Breath Volume	1.335	03:25
Air Blank	0.000	03:25
Air Blank	0.000	03:27
Subject Sample # 2	0.193	03:28
Breath Volume	2.460	03:28
Air Blank	0.000	03:28
Control Test	0.076	03:29
Air Blank	0.000	03:29
Diagnostics Check	OK	03:29

---

Data Download Date/Time: 10/22/2016 21:31

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006629

INSTRUMENT REGISTERED TO: FHP  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HAWKINS,DEBORAH,  
BREATH TEST OPERATOR AGENCY: FHP Troop D  
ARRESTING OFFICER: HAWKINS,DEBORAH  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 09/30/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 09216080A3  
DRY GAS EXPIRY DATE: 05/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 03:16  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: VACHA

DATE/TIME OF TEST: 10/13/2016 03:58:00

---

Diagnostics Check	OK	03:50
Air Blank	0.000	03:51
Control Test	0.078	03:51
Air Blank	0.000	03:52
Subject Sample # 1	0.155	03:53
Breath Volume	2.886	03:53
Air Blank	0.000	03:53
Air Blank	0.000	03:55
Subject Sample # 2	0.159	03:56
Breath Volume	2.988	03:56
Air Blank	0.000	03:57
Control Test	0.077	03:57
Air Blank	0.000	03:57
Diagnostics Check	OK	03:58

---

Data Download Date/Time: 10/22/2016 21:31



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006629

INSTRUMENT REGISTERED TO: FHP  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HAWKINS,DEBORAH,  
BREATH TEST OPERATOR AGENCY: FHP Troop D  
ARRESTING OFFICER: HAWKINS,DEBORAH  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 09/30/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 09216080A3  
DRY GAS EXPIRY DATE: 05/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 03:17  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:41

SUBJECT'S LAST NAME: FINGAR

DATE/TIME OF TEST: 10/16/2016 04:12:50

---

Diagnostics Check	OK	04:06
Air Blank	0.000	04:06
Control Test	0.080	04:07
Air Blank	0.000	04:07
Subject Sample # 1	0.123	04:08
Breath Volume	1.671	04:08
Air Blank	0.000	04:08
Air Blank	0.000	04:10
Subject Sample # 2	0.128	04:11
Breath Volume	3.277	04:11
Air Blank	0.000	04:11
Control Test	0.078	04:12
Air Blank	0.000	04:12
Diagnostics Check	OK	04:12

---

Data Download Date/Time: 10/22/2016 21:31

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006629

INSTRUMENT REGISTERED TO: FHP  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HAWKINS,DEBORAH,  
BREATH TEST OPERATOR AGENCY: FHP Troop D  
ARRESTING OFFICER: HAWKINS,DEBORAH  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 09/30/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 09216080A3  
DRY GAS EXPIRY DATE: 05/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 02:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 10/19/2016 03:11:14

---

Diagnostics Check	OK	02:58	
Air Blank	0.000	02:58	
Control Test	0.079	02:58	
Air Blank	0.000	02:59	
Subject Sample	# 1 0.080	03:00	
Breath Volume	1.140	03:00	
Air Blank	0.000	03:01	
Air Blank	0.000	03:03	
Subject Sample	# 2 0.072	03:06	Volume Not Met
Breath Volume	0.082	03:06	
Air Blank	0.000	03:07	
Air Blank	0.000	03:09	
Subject Sample	# 3 0.083	03:09	
Breath Volume	2.484	03:09	
Air Blank	0.000	03:10	
Control Test	0.079	03:10	
Air Blank	0.000	03:11	
Diagnostics Check	OK	03:11	

---

Data Download Date/Time: 10/22/2016 21:31

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006629

INSTRUMENT REGISTERED TO: FHP  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HAWKINS,DEBORAH,  
BREATH TEST OPERATOR AGENCY: FHP Troop D  
ARRESTING OFFICER: HAWKINS,DEBORAH  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 10/22/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 09216080A3  
DRY GAS EXPIRY DATE: 05/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 22:25  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:01

SUBJECT'S LAST NAME: ARMAS

DATE/TIME OF TEST: 10/25/2016 23:33:03

---

Diagnostics Check	OK	23:26
Air Blank	0.000	23:27
Control Test	0.080	23:27
Air Blank	0.000	23:28
Subject Sample # 1	0.101	23:28
Breath Volume	2.539	23:28
Air Blank	0.000	23:29
Air Blank	0.000	23:31
Subject Sample # 2	0.097	23:31
Breath Volume	2.339	23:31
Air Blank	0.000	23:32
Control Test	0.079	23:32
Air Blank	0.000	23:32
Diagnostics Check	OK	23:33

---

Data Download Date/Time: 11/30/2016 22:09

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006629

INSTRUMENT REGISTERED TO: FHP  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HAWKINS,DEBORAH,  
BREATH TEST OPERATOR AGENCY: FHP Troop D  
ARRESTING OFFICER: HAWKINS,DEBORAH'  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 10/22/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 09216080A3  
DRY GAS EXPIRY DATE: 05/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 03:37  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: CREECH

DATE/TIME OF TEST: 10/26/2016 04:16:19

---

Diagnostics Check	OK	04:10
Air Blank	0.000	04:10
Control Test	0.079	04:10
Air Blank	0.000	04:11
Subject Sample # 1	0.167	04:11
Breath Volume	3.097	04:11
Air Blank	0.000	04:12
Air Blank	0.000	04:14
Subject Sample # 2	0.166	04:14
Breath Volume	2.035	04:14
Air Blank	0.000	04:15
Control Test	0.079	04:15
Air Blank	0.000	04:16
Diagnostics Check	OK	04:16

---

Data Download Date/Time: 11/30/2016 22:09

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006630

INSTRUMENT REGISTERED TO: FHP  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GONZALEZ-CORTINA,RAUL,H  
BREATH TEST OPERATOR AGENCY: FHP Troop E  
ARRESTING OFFICER: GONZALEZ-CORTINA,RAUL  
ARRESTING OFFICER AGENCY: FHP Troop E  
LAST AGENCY INSPECTION: 09/16/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06616080A2  
DRY GAS EXPIRY DATE: 04/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 03:13  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:13

SUBJECT'S LAST NAME: BIAMBY

DATE/TIME OF TEST: 10/22/2016 04:57:32

---

Diagnostics Check	OK	04:53	
Air Blank	0.000	04:54	
Control Test	0.079	04:54	
Air Blank	0.000	04:55	
Subject Sample	# 1	04:56	RFI Detect
Breath Volume	0.000	04:56	
Air Blank	0.000	04:56	
Control Test	0.079	04:56	
Air Blank	0.000	04:57	
Diagnostics Check	OK	04:57	

---

Data Download Date/Time: 10/31/2016 08:59

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006630

INSTRUMENT REGISTERED TO: FHP  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GONZALEZ-CORTINA,RAUL,H  
BREATH TEST OPERATOR AGENCY: FHP Troop E  
ARRESTING OFFICER: GONZALEZ-CORTINA,RAUL  
ARRESTING OFFICER AGENCY: FHP Troop E  
LAST AGENCY INSPECTION: 09/16/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06616080A2  
DRY GAS EXPIRY DATE: 04/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 03:13  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:13

SUBJECT'S LAST NAME: BIAMBY

DATE/TIME OF TEST: 10/22/2016 05:03:09

---

Diagnostics Check	OK	05:00	
Air Blank	0.000	05:00	
Control Test	0.078	05:00	
Air Blank	0.000	05:01	
Subject Sample	# 1	05:01	Subject Test Refused
Breath Volume	0.000	05:01	
Air Blank	0.000	05:02	
Control Test	0.078	05:02	
Air Blank	0.000	05:03	
Diagnostics Check	OK	05:03	

---

Data Download Date/Time: 10/31/2016 08:59

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006630

INSTRUMENT REGISTERED TO: FHP  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GONZALEZ-CORTINA,RAUL,H  
BREATH TEST OPERATOR AGENCY: FHP Troop E  
ARRESTING OFFICER: GONZALEZ-CORTINA,RAUL  
ARRESTING OFFICER AGENCY: FHP Troop E  
LAST AGENCY INSPECTION: 09/16/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06616080A2  
DRY GAS EXPIRY DATE: 04/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 03:58  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: LEDO-ALVAREZ

DATE/TIME OF TEST: 10/27/2016 05:36:41

---

Diagnostics Check	OK	05:29
Air Blank	0.000	05:30
Control Test	0.079	05:30
Air Blank	0.000	05:30
Subject Sample # 1	0.143	05:32
Breath Volume	1.652	05:32
Air Blank	0.000	05:32
Air Blank	0.000	05:34
Subject Sample # 2	0.159	05:35
Breath Volume	2.132	05:35
Air Blank	0.000	05:35
Control Test	0.078	05:36
Air Blank	0.000	05:36
Diagnostics Check	OK	05:36

---

Data Download Date/Time: 10/31/2016 08:59

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006630

INSTRUMENT REGISTERED TO: FHP  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GONZALEZ-CORTINA,RAUL,H  
BREATH TEST OPERATOR AGENCY: FHP Troop E  
ARRESTING OFFICER: GONZALEZ-CORTINA,RAUL  
ARRESTING OFFICER AGENCY: FHP Troop E  
LAST AGENCY INSPECTION: 09/16/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06616080A2  
DRY GAS EXPIRY DATE: 04/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 02:27  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:27

SUBJECT'S LAST NAME: INGRAM

DATE/TIME OF TEST: 10/29/2016 04:15:39

---

Diagnostics Check		OK	04:00	
Air Blank		0.000	04:01	
Control Test		0.079	04:01	
Air Blank		0.000	04:01	
Subject Sample	# 1	0.107	04:04	Volume Not Met
Breath Volume		0.035	04:04	
Air Blank		0.000	04:05	
Air Blank		0.000	04:07	
Subject Sample	# 2	0.115	04:08	
Breath Volume		1.785	04:08	
Air Blank		0.000	04:09	
Air Blank		0.000	04:10	
Subject Sample	# 3	0.102	04:14	Volume Not Met
Breath Volume		0.820	04:14	
Air Blank		0.000	04:14	
Control Test		0.079	04:15	
Air Blank		0.000	04:15	
Diagnostics Check		OK	04:15	

---

Data Download Date/Time: 10/31/2016 08:59



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006630

INSTRUMENT REGISTERED TO: FHP  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GONZALEZ-CORTINA,RAUL,H  
BREATH TEST OPERATOR AGENCY: FHP Troop E  
ARRESTING OFFICER: GONZALEZ-CORTINA,RAUL  
ARRESTING OFFICER AGENCY: FHP Troop E  
LAST AGENCY INSPECTION: 09/16/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06616080A2  
DRY GAS EXPIRY DATE: 04/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 02:27  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:27

SUBJECT'S LAST NAME: INGRAM

DATE/TIME OF TEST: 10/29/2016 04:25:06

---

Diagnostics Check		OK	04:17
Air Blank		0.000	04:17
Control Test		0.077	04:18
Air Blank		0.000	04:18
Subject Sample	# 1	0.115	04:20
Breath Volume		1.347	04:20
Air Blank		0.000	04:21
Air Blank		0.000	04:23
Subject Sample	# 2	0.116	04:23
Breath Volume		2.265	04:23
Air Blank		0.000	04:24
Control Test		0.077	04:24
Air Blank		0.000	04:25
Diagnostics Check		OK	04:25

---

Data Download Date/Time: 10/31/2016 08:59

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006630

INSTRUMENT REGISTERED TO: FHP  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GONZALEZ-CORTINA,RAUL,H  
BREATH TEST OPERATOR AGENCY: FHP Troop E  
ARRESTING OFFICER: LLANIO,PIERRE  
ARRESTING OFFICER AGENCY: FHP Troop E  
LAST AGENCY INSPECTION: 09/16/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06616080A2  
DRY GAS EXPIRY DATE: 04/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 02:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: FAJARDO GOMEZ

DATE/TIME OF TEST: 10/30/2016 03:46:46

---

Diagnostics Check	OK	03:40
Air Blank	0.000	03:40
Control Test	0.079	03:40
Air Blank	0.000	03:41
Subject Sample # 1	0.165	03:42
Breath Volume	3.144	03:42
Air Blank	0.000	03:42
Air Blank	0.000	03:44
Subject Sample # 2	0.162	03:45
Breath Volume	3.203	03:45
Air Blank	0.000	03:45
Control Test	0.078	03:46
Air Blank	0.000	03:46
Diagnostics Check	OK	03:46

---

Data Download Date/Time: 10/31/2016 08:59

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006630

INSTRUMENT REGISTERED TO: FHP  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GONZALEZ-CORTINA,RAUL,H  
BREATH TEST OPERATOR AGENCY: FHP Troop E  
ARRESTING OFFICER: GONZALEZ-CORTINA,RAUL  
ARRESTING OFFICER AGENCY: FHP Troop E  
LAST AGENCY INSPECTION: 09/16/2016

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 06616080A2  
DRY GAS EXPIRY DATE: 04/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 03:06  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:06

SUBJECT'S LAST NAME: VALVERDE

DATE/TIME OF TEST: 10/30/2016 03:50:54

---

Diagnostics Check	OK	03:44
Air Blank	0.000	03:44
Control Test	0.079	03:45
Air Blank	0.000	03:45
Subject Sample # 1	0.204	03:46
Breath Volume	2.261	03:46
Air Blank	0.000	03:47
Air Blank	0.000	03:48
Subject Sample # 2	0.197	03:49
Breath Volume	1.718	03:49
Air Blank	0.000	03:49
Control Test	0.077	03:50
Air Blank	0.000	03:50
Diagnostics Check	OK	03:50

---

Data Download Date/Time: 10/31/2016 08:59

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006633

INSTRUMENT REGISTERED TO: FHP  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ANGELICCHI,JOSEPH,S  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: ANGELICCHI,JOSEPH  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 09/28/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 09216080A3  
DRY GAS EXPIRY DATE: 05/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 01:04  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: BENNETT

DATE/TIME OF TEST: 10/01/2016 01:16:36

---

Diagnostics Check	OK	01:13	
Air Blank	0.000	01:14	
Control Test	0.078	01:14	
Air Blank	0.000	01:15	
Subject Sample	# 1	01:15	Subject Test Refused
Breath Volume	0.000	01:15	
Air Blank	0.000	01:15	
Control Test	0.077	01:15	
Air Blank	0.000	01:16	
Diagnostics Check	OK	01:16	

---

Data Download Date/Time: 10/26/2016 22:05

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006633

INSTRUMENT REGISTERED TO: FHP  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ANGELICCHI,JOSEPH,S  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: ANGELICCHI,JOSEPH  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 09/28/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 09216080A3  
DRY GAS EXPIRY DATE: 05/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 02:09  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: BOTE

DATE/TIME OF TEST: 10/02/2016 02:12:53

---

Diagnostics Check	OK	02:12	
Air Blank		02:12	RFI Detect
Air Blank		02:12	RFI Detect

---

Data Download Date/Time: 10/26/2016 22:05

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006633

INSTRUMENT REGISTERED TO: FHP  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ANGELICCHI,JOSEPH,S  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: ANGELICCHI,JOSEPH  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 09/28/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 09216080A3  
DRY GAS EXPIRY DATE: 05/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 06:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:20

SUBJECT'S LAST NAME: ASKAR

DATE/TIME OF TEST: 10/22/2016 06:50:17

---

Diagnostics Check	OK	06:47	
Air Blank	0.000	06:47	
Control Test	0.078	06:48	
Air Blank	0.000	06:48	
Subject Sample	# 1	06:48	Subject Test Refused
Breath Volume	0.000	06:48	
Air Blank	0.000	06:49	
Control Test	0.079	06:49	
Air Blank	0.000	06:50	
Diagnostics Check	OK	06:50	

---

Data Download Date/Time: 10/26/2016 22:05

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006633

INSTRUMENT REGISTERED TO: FHP  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ANGELICCHI,JOSEPH,S  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: ANGELICCHI,JOSEPH  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 09/28/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 09216080A3  
DRY GAS EXPIRY DATE: 05/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 20:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:40

SUBJECT'S LAST NAME: ROBB

DATE/TIME OF TEST: 10/25/2016 21:11:09

---

Diagnostics Check	OK	21:04
Air Blank	0.000	21:05
Control Test	0.079	21:05
Air Blank	0.000	21:06
Subject Sample # 1	0.141	21:06
Breath Volume	2.675	21:06
Air Blank	0.000	21:07
Air Blank	0.000	21:08
Subject Sample # 2	0.148	21:09
Breath Volume	2.796	21:09
Air Blank	0.000	21:10
Control Test	0.079	21:10
Air Blank	0.000	21:11
Diagnostics Check	OK	21:11

---

Data Download Date/Time: 10/26/2016 22:05

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: VANOVER,ANDREW,P  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: VANOVER,ANDREW  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 09/26/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 256835  
DRY GAS EXPIRY DATE: 01/11/2019  
VIOLATION CODE: DUI  
ARREST TIME: 18:24  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:32

SUBJECT'S LAST NAME: HARTMAN

DATE/TIME OF TEST: 10/08/2016 18:59:07

---

Diagnostics Check	OK	18:52
Air Blank	0.000	18:53
Control Test	0.080	18:53
Air Blank	0.000	18:54
Subject Sample # 1	0.014	18:54
Breath Volume	1.660	18:54
Air Blank	0.000	18:55
Air Blank	0.000	18:57
Subject Sample # 2	0.011	18:57
Breath Volume	2.093	18:57
Air Blank	0.000	18:58
Control Test	0.078	18:58
Air Blank	0.000	18:59
Diagnostics Check	OK	18:59

---

Data Download Date/Time: 10/17/2016 10:47



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TATE,ROBERT,J  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: TATE,ROBERT  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 09/26/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 256835  
DRY GAS EXPIRY DATE: 01/11/2019  
VIOLATION CODE: DUI  
ARREST TIME: 21:26  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: BAKER

DATE/TIME OF TEST: 10/15/2016 22:33:52

---

Diagnostics Check	OK	22:30	
Air Blank	0.000	22:30	
Control Test	0.080	22:31	
Air Blank	0.000	22:31	
Subject Sample	# 1	22:32	Subject Test Refused
Breath Volume	0.000	22:32	
Air Blank	0.000	22:32	
Control Test	0.081	22:33	
Air Blank	0.000	22:33	
Diagnostics Check	OK	22:33	

---

Data Download Date/Time: 10/17/2016 10:47

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: VANOVER,ANDREW,P  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: VANOVER,ANDREW  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 10/17/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 256835  
DRY GAS EXPIRY DATE: 01/11/2019  
VIOLATION CODE: DUI  
ARREST TIME: 17:01  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:27

SUBJECT'S LAST NAME: DELANGIS

DATE/TIME OF TEST: 10/27/2016 17:54:09

---

Diagnostics Check	OK	17:48
Air Blank	0.000	17:48
Control Test	0.081	17:48
Air Blank	0.000	17:49
Subject Sample # 1	0.000	17:49
Breath Volume	4.062	17:49
Air Blank	0.000	17:50
Air Blank	0.000	17:52
Subject Sample # 2	0.000	17:52
Breath Volume	3.847	17:52
Air Blank	0.000	17:53
Control Test	0.079	17:53
Air Blank	0.000	17:54
Diagnostics Check	OK	17:54

---

Data Download Date/Time: 11/28/2016 09:25

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006635

INSTRUMENT REGISTERED TO: FHP  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MOONDA,SHUAIB,  
BREATH TEST OPERATOR AGENCY: FHP Troop D  
ARRESTING OFFICER: JORGE,JUNIOR  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 09/30/2016

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 3RTG  
DRY GAS EXPIRY DATE: 11/11/2022  
VIOLATION CODE: DUI  
ARREST TIME: 19:03  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 19:05

SUBJECT'S LAST NAME: WEEKS

DATE/TIME OF TEST: 10/07/2016 19:35:39

---

Diagnostics Check	OK	19:27
Air Blank	0.000	19:28
Control Test	0.077	19:28
Air Blank	0.000	19:28
Subject Sample # 1	0.000	19:30
Breath Volume	1.257	19:30
Air Blank	0.000	19:31
Air Blank	0.000	19:33
Subject Sample # 2	0.000	19:34
Breath Volume	2.726	19:34
Air Blank	0.000	19:34
Control Test	0.078	19:35
Air Blank	0.000	19:35
Diagnostics Check	OK	19:35

---

Data Download Date/Time: 10/22/2016 21:34

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006635

INSTRUMENT REGISTERED TO: FHP  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MOONDA,SHUAIB,  
BREATH TEST OPERATOR AGENCY: FHP Troop D  
ARRESTING OFFICER: BINET,MARLON  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 09/30/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 3RTG  
DRY GAS EXPIRY DATE: 05/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 00:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: ROSA SANTIAGO

DATE/TIME OF TEST: 10/09/2016 01:51:46

---

Diagnostics Check	OK	01:45
Air Blank	0.000	01:45
Control Test	0.079	01:46
Air Blank	0.000	01:46
Subject Sample # 1	0.126	01:47
Breath Volume	3.617	01:47
Air Blank	0.000	01:47
Air Blank	0.000	01:49
Subject Sample # 2	0.124	01:50
Breath Volume	3.746	01:50
Air Blank	0.000	01:50
Control Test	0.078	01:51
Air Blank	0.000	01:51
Diagnostics Check	OK	01:51

---

Data Download Date/Time: 10/22/2016 21:34

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006635

INSTRUMENT REGISTERED TO: FHP

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MOONDA,SHUAIB,

BREATH TEST OPERATOR AGENCY: FHP Troop D

ARRESTING OFFICER: IV,ANDREW

ARRESTING OFFICER AGENCY: FHP Troop D

LAST AGENCY INSPECTION: 09/30/2016

Date or Time Changed: No

DRY GAS STANDARD LOT #: 3RTG

DRY GAS EXPIRY DATE: 09/05/2018

VIOLATION CODE: DUI

ARREST TIME: 03:28

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: NEGRON

DATE/TIME OF TEST: 10/18/2016 04:03:54

---

Diagnostics Check	OK	04:02	
Air Blank	0.000	04:02	
Control Test	0.078	04:02	
Air Blank	0.000	04:03	
		04:03	RFI Detect
Air Blank	0.000	04:03	

---

Data Download Date/Time: 10/22/2016 21:34

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006635

INSTRUMENT REGISTERED TO: FHP  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MOONDA,SHUAIB,  
BREATH TEST OPERATOR AGENCY: FHP Troop D  
ARRESTING OFFICER: IV,ANDREW  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 09/30/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 3RTG  
DRY GAS EXPIRY DATE: 09/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 03:28  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: NEGRON

DATE/TIME OF TEST: 10/18/2016 04:12:38

---

Diagnostics Check	OK	04:06
Air Blank	0.000	04:06
Control Test	0.079	04:07
Air Blank	0.000	04:07
Subject Sample # 1	0.000	04:07
Breath Volume	3.910	04:07
Air Blank	0.000	04:08
Air Blank	0.000	04:10
Subject Sample # 2	0.000	04:11
Breath Volume	3.820	04:11
Air Blank	0.000	04:11
Control Test	0.079	04:11
Air Blank	0.000	04:12
Diagnostics Check	OK	04:12

---

Data Download Date/Time: 10/22/2016 21:34

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006635

INSTRUMENT REGISTERED TO: FHP  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MOONDA,SHUAIB,  
BREATH TEST OPERATOR AGENCY: FHP Troop D  
ARRESTING OFFICER: MOONDA,SHUAIB  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 09/30/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 3RTG  
DRY GAS EXPIRY DATE: 09/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 01:41  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: MONTILLA TRUJILLO

DATE/TIME OF TEST: 10/21/2016 02:18:40

---

Diagnostics Check	OK	02:11
Air Blank	0.000	02:11
Control Test	0.079	02:11
Air Blank	0.000	02:12
Subject Sample # 1	0.122	02:12
Breath Volume	3.121	02:12
Air Blank	0.000	02:13
Air Blank	0.000	02:15
Subject Sample # 2	0.127	02:17
Breath Volume	2.871	02:17
Air Blank	0.000	02:17
Control Test	0.079	02:18
Air Blank	0.000	02:18
Diagnostics Check	OK	02:18

---

Data Download Date/Time: 10/22/2016 21:34

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006636

INSTRUMENT REGISTERED TO: FHP  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HITCHCOCK, WILLIAM, D  
BREATH TEST OPERATOR AGENCY: FHP Troop D  
ARRESTING OFFICER: HITCHCOCK, WILLIAM  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 09/30/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 09216080A3  
DRY GAS EXPIRY DATE: 05/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 01:36  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:23

SUBJECT'S LAST NAME: DAWKINS

DATE/TIME OF TEST: 10/21/2016 03:10:48

---

Diagnostics Check	OK	02:59	
Air Blank	0.000	02:59	
Control Test	0.080	03:00	
Air Blank	0.000	03:00	
Subject Sample	# 1	03:03	No Sample Provided
Breath Volume	0.000	03:03	
Air Blank	0.000	03:04	
Air Blank	0.000	03:06	
Subject Sample	# 2	03:09	No Sample Provided
Breath Volume	0.000	03:09	
Air Blank	0.000	03:09	
Control Test	0.080	03:10	
Air Blank	0.000	03:10	
Diagnostics Check	OK	03:10	

---

Data Download Date/Time: 10/27/2016 00:00



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006636

INSTRUMENT REGISTERED TO: FHP  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HITCHCOCK, WILLIAM, D  
BREATH TEST OPERATOR AGENCY: FHP Troop D  
ARRESTING OFFICER: HITCHCOCK, WILLIAM  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 09/30/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 09216080A3  
DRY GAS EXPIRY DATE: 05/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 23:39  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:21

SUBJECT'S LAST NAME: KISSINGER JR

DATE/TIME OF TEST: 10/26/2016 00:55:33

---

Diagnostics Check	OK	00:44	
Air Blank	0.000	00:44	
Control Test	0.080	00:44	
Air Blank	0.000	00:45	
Subject Sample	# 1	00:48	No Sample Provided
Breath Volume	0.000	00:48	
Air Blank	0.000	00:49	
Air Blank	0.000	00:51	
Subject Sample	# 2	00:54	No Sample Provided
Breath Volume	0.000	00:54	
Air Blank	0.000	00:54	
Control Test	0.080	00:54	
Air Blank	0.000	00:55	
Diagnostics Check	OK	00:55	

---

Data Download Date/Time: 10/27/2016 00:00

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006692

INSTRUMENT REGISTERED TO: HOLMES BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BORES,ALAN,P  
BREATH TEST OPERATOR AGENCY: Holmes Beach PD  
ARRESTING OFFICER: BORES,ALAN  
ARRESTING OFFICER AGENCY: Holmes Beach PD  
LAST AGENCY INSPECTION: 10/17/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17316080A2  
DRY GAS EXPIRY DATE: 08/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 01:22  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:38

SUBJECT'S LAST NAME: BRAATEN

DATE/TIME OF TEST: 10/29/2016 02:05:34

---

Diagnostics Check	OK	02:02	
Air Blank	0.000	02:03	
Control Test	0.078	02:03	
Air Blank	0.000	02:04	
Subject Sample	# 1	02:04	Subject Test Refused
Breath Volume	0.000	02:04	
Air Blank	0.000	02:04	
Control Test	0.078	02:04	
Air Blank	0.000	02:05	
Diagnostics Check	OK	02:05	

---

Data Download Date/Time: 11/28/2016 10:35

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006692

INSTRUMENT REGISTERED TO: HOLMES BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BORES,ALAN,P  
BREATH TEST OPERATOR AGENCY: Holmes Beach PD  
ARRESTING OFFICER: BORES,ALAN  
ARRESTING OFFICER AGENCY: Holmes Beach PD  
LAST AGENCY INSPECTION: 10/17/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17316080A2  
DRY GAS EXPIRY DATE: 08/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 01:22  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:38

SUBJECT'S LAST NAME: BRAATEN

DATE/TIME OF TEST: 10/29/2016 02:12:33

---

Diagnostics Check	OK	02:09	
Air Blank	0.000	02:10	
Control Test	0.078	02:10	
Air Blank	0.000	02:11	
Subject Sample	# 1	02:11	Subject Test Refused
Breath Volume	0.000	02:11	
Air Blank	0.000	02:11	
Control Test	0.078	02:11	
Air Blank	0.000	02:12	
Diagnostics Check	OK	02:12	

---

Data Download Date/Time: 11/28/2016 10:35