Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,B,
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: HOLLCROFT,D
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 12/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A2
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:17

SUBJECT'S LAST NAME: ETHERIDGE

DATE/TIME OF TEST: 01/01/2017 03:44:23

Diagnostics Check		OK	03:37
Air Blank		0.000	03:38
Control Test		0.080	03:38
Air Blank		0.000	03:39
Subject Sample	# 1	0.076	03:39
Breath Volume		1.441	03:39
Air Blank		0.000	03:40
Air Blank		0.000	03:42
Subject Sample	# 2	0.071	03:42
Breath Volume		2.632	03:42
Air Blank		0.000	03:43
Control Test		0.077	03:43
Air Blank		0.000	03:44
Diagnostics Check		OK	03:44

Data Download Date/Time: 01/26/2017 14:06

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,B,
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: HOLLCROFT,D
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 12/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A2
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 05:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:02

SUBJECT'S LAST NAME: RODGERS

DATE/TIME OF TEST: 01/01/2017 05:43:52

Diagnostics Check		OK	05:36
Air Blank		0.000	05:37
Control Test		0.079	05:37
Air Blank		0.000	05:38
Subject Sample	# 1	0.043	05:39
Breath Volume		1.855	05:39
Air Blank		0.000	05:39
Air Blank		0.000	05:41
Subject Sample	# 2	0.041	05:42
Breath Volume		2.082	05:42
Air Blank		0.000	05:42
Control Test		0.078	05:43
Air Blank		0.000	05:43
Diagnostics Check		OK	05:43

Data Download Date/Time: 01/26/2017 14:06

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIDGEON,ANDREW,W
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: VURRO,MARC
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 12/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A2
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 15:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:05

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 01/08/2017 15:47:07

Diagnostics Check		OK	15:39
Air Blank		0.000	15:40
Control Test		0.079	15:40
Air Blank		0.000	15:41
Subject Sample	# 1	0.139	15:42
Breath Volume		1.292	15:42
Air Blank		0.000	15:43
Air Blank		0.000	15:45
Subject Sample	# 2	0.141	15:45
Breath Volume		1.574	15:45
Air Blank		0.000	15:46
Control Test		0.077	15:46
Air Blank		0.000	15:47
Diagnostics Check		OK	15:47

Data Download Date/Time: 01/26/2017 14:06

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FREY,ROBERT,P
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: ROLPH,A
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 12/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A2
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: HALE

DATE/TIME OF TEST: 01/25/2017 00:30:45

Diagnostics Check		OK	00:24
Air Blank		0.000	00:25
Control Test		0.079	00:25
Air Blank		0.000	00:25
Subject Sample	# 1	0.133	00:26
Breath Volume		2.339	00:26
Air Blank		0.000	00:27
Air Blank		0.000	00:28
Subject Sample	# 2	0.136	00:29
Breath Volume		2.015	00:29
Air Blank		0.000	00:29
Control Test		0.080	00:30
Air Blank		0.000	00:30
Diagnostics Check		OK	00:30

Data Download Date/Time: 01/26/2017 14:06

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JASIUKIEWICZ,MICHAEL,P
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: JASIUKIEWICZ,MICHAEL
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 12/08/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: SOK

DATE/TIME OF TEST: 01/01/2017 01:41:31

Diagnostics Check		OK	01:35
Air Blank		0.000	01:36
Control Test		0.080	01:36
Air Blank		0.000	01:36
Subject Sample	# 1	0.000	01:37
Breath Volume		2.953	01:37
Air Blank		0.000	01:37
Air Blank		0.000	01:39
Subject Sample	# 2	0.000	01:40
Breath Volume		2.808	01:40
Air Blank		0.000	01:40
Control Test		0.080	01:40
Air Blank		0.000	01:41
Diagnostics Check		OK	01:41

Data Download Date/Time: 01/12/2017 08:22

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROLLESTON,JACOB,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: ROLLESTON,JACOB
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 12/08/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: NADEAU

DATE/TIME OF TEST: 01/01/2017 02:27:13

Diagnostics Check		OK	02:20
Air Blank		0.000	02:21
Control Test		0.079	02:21
Air Blank		0.000	02:22
Subject Sample	# 1	0.155	02:22
Breath Volume		1.398	02:22
Air Blank		0.000	02:23
Air Blank		0.000	02:24
Subject Sample	# 2	0.156	02:25
Breath Volume		1.527	02:25
Air Blank		0.000	02:26
Control Test		0.078	02:26
Air Blank		0.000	02:27
Diagnostics Check		OK	02:27

Data Download Date/Time: 01/12/2017 08:22

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROLLESTON,JACOB,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: ROLLESTON,JACOB
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 12/08/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: CRANER

DATE/TIME OF TEST: 01/06/2017 02:14:24

Diagnostics Check		OK	02:05
Air Blank		0.000	02:06
Control Test		0.078	02:06
Air Blank		0.000	02:07
Subject Sample	# 1	0.191	02:09
Breath Volume		1.531	02:09
Air Blank		0.000	02:10
Air Blank		0.000	02:11
Subject Sample	# 2	0.179	02:12
Breath Volume		1.691	02:12
Air Blank		0.000	02:13
Control Test		0.075	02:13
Air Blank		0.000	02:14
Diagnostics Check		OK	02:14

Data Download Date/Time: 01/12/2017 08:22

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JASIUKIEWICZ,MICHAEL,P
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: JASIUKIEWICZ,MICHAEL
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 12/08/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: SEITZ

DATE/TIME OF TEST: 01/06/2017 02:58:54

Diagnostics Check		OK	02:56	
Air Blank		0.000	02:56	
Control Test		0.077	02:56	
Air Blank		0.000	02:57	
Subject Sample	# 1		02:57	Subject Test Refused
Breath Volume		0.000	02:57	
Air Blank		0.000	02:57	
Control Test		0.078	02:58	
Air Blank		0.000	02:58	
Diagnostics Check		OK	02:58	

Data Download Date/Time: 01/12/2017 08:22

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALLEY,CPL,S
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: GALLEY,CPL
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 12/08/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 13:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:18

SUBJECT'S LAST NAME: WESTON

DATE/TIME OF TEST: 01/06/2017 14:48:56

Diagnostics Check		OK	14:38
Air Blank		0.000	14:38
Control Test		0.079	14:39
Air Blank		0.000	14:39
Subject Sample	# 1	0.215	14:40
Breath Volume		2.484	14:40
Air Blank		0.000	14:40
Air Blank		0.000	14:42
Subject Sample	# 2	0.174	14:43
Breath Volume		1.179	14:43
Air Blank		0.000	14:43
Air Blank		0.000	14:45
Subject Sample	# 3	0.171	14:47
Breath Volume		1.472	14:47
Air Blank		0.000	14:47
Control Test		0.077	14:48
Air Blank		0.000	14:48
Diagnostics Check		OK	14:48

Data Download Date/Time: 01/12/2017 08:22

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JASIUKIEWICZ,MICHAEL,P
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: JASIUKIEWICZ,MICHAEL
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 12/08/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:24

SUBJECT'S LAST NAME: LAGOTTE

DATE/TIME OF TEST: 01/06/2017 20:48:55

Diagnostics Check		OK	20:45	
Air Blank		0.000	20:46	
Control Test		0.078	20:46	
Air Blank		0.000	20:47	
Subject Sample	# 1		20:47	Subject Test Refused
Breath Volume		0.000	20:47	
Air Blank		0.000	20:48	
Control Test		0.078	20:48	
Air Blank		0.000	20:48	
Diagnostics Check		OK	20:48	

Data Download Date/Time: 01/12/2017 08:22

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JASIUKIEWICZ,MICHAEL,P
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: JASIUKIEWICZ,MICHAEL
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 12/08/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:58

SUBJECT'S LAST NAME: SOLKA

DATE/TIME OF TEST: 01/07/2017 03:24:14

Diagnostics Check		OK	03:18
Air Blank		0.000	03:18
Control Test		0.078	03:18
Air Blank		0.000	03:19
Subject Sample	# 1	0.151	03:19
Breath Volume		4.011	03:19
Air Blank		0.000	03:20
Air Blank		0.000	03:22
Subject Sample	# 2	0.142	03:22
Breath Volume		3.941	03:22
Air Blank		0.000	03:23
Control Test		0.078	03:23
Air Blank		0.000	03:24
Diagnostics Check		OK	03:24

Data Download Date/Time: 01/12/2017 08:22

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JASIUKIEWICZ,MICHAEL,P
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: JASIUKIEWICZ,MICHAEL
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 12/08/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:54

SUBJECT'S LAST NAME: OLES

DATE/TIME OF TEST: 01/08/2017 01:20:57

Diagnostics Check		OK	01:14
Air Blank		0.000	01:15
Control Test		0.080	01:15
Air Blank		0.000	01:16
Subject Sample	# 1	0.139	01:16
Breath Volume		2.902	01:16
Air Blank		0.000	01:17
Air Blank		0.000	01:18
Subject Sample	# 2	0.141	01:19
Breath Volume		2.457	01:19
Air Blank		0.000	01:20
Control Test		0.079	01:20
Air Blank		0.000	01:20
Diagnostics Check		OK	01:20

Data Download Date/Time: 01/12/2017 08:22

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOYLE,THOMAS,E
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: BOYLE,THOMAS
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 12/08/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 19:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:55

SUBJECT'S LAST NAME: HAGER

DATE/TIME OF TEST: 01/08/2017 20:21:56

Diagnostics Check		OK	20:17	
Air Blank		0.000	20:17	
Control Test		0.080	20:18	
Air Blank		0.000	20:18	
Subject Sample	# 1		20:20	Subject Test Refused
Breath Volume		0.000	20:20	
Air Blank		0.000	20:21	
Control Test		0.079	20:21	
Air Blank		0.000	20:21	
Diagnostics Check		OK	20:21	

Data Download Date/Time: 01/12/2017 08:22

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHEA,CPL,J
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: SHEA,JOHN
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 12/08/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:31

SUBJECT'S LAST NAME: URQUHART

DATE/TIME OF TEST: 01/08/2017 23:02:03

Diagnostics Check		OK	22:53	
Air Blank		0.000	22:54	
Control Test		0.080	22:54	
Air Blank		0.000	22:54	
Subject Sample	# 1		22:58	No Sample Provided
Breath Volume		0.000	22:58	
Air Blank		0.000	22:58	
Air Blank		0.000	23:00	
Subject Sample	# 2		23:00	Subject Test Refused
Breath Volume		0.000	23:00	
Air Blank		0.000	23:01	
Control Test		0.080	23:01	
Air Blank		0.000	23:01	
Diagnostics Check		OK	23:02	

Data Download Date/Time: 01/12/2017 08:22

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHEA,CPL,J
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: SHEA,JOHN
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 12/08/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: BISTOFF

DATE/TIME OF TEST: 01/10/2017 03:09:11

Diagnostics Check		OK	03:06	
Air Blank		0.000	03:06	
Control Test		0.080	03:07	
Air Blank		0.000	03:07	
Subject Sample	# 1		03:07	Subject Test Refused
Breath Volume		0.000	03:07	
Air Blank		0.000	03:08	
Control Test		0.080	03:08	
Air Blank		0.000	03:09	
Diagnostics Check		OK	03:09	

Data Download Date/Time: 01/12/2017 08:22

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHEA,CPL,J
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: MORGAN,WILLIAM
ARRESTING OFFICER AGENCY: Kenneth City PD
LAST AGENCY INSPECTION: 12/08/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:46

SUBJECT'S LAST NAME: RISCH

DATE/TIME OF TEST: 01/11/2017 02:13:12

Diagnostics Check		OK	02:07
Air Blank		0.000	02:07
Control Test		0.080	02:07
Air Blank		0.000	02:08
Subject Sample	# 1	0.000	02:08
Breath Volume		3.042	02:08
Air Blank		0.000	02:09
Air Blank		0.000	02:11
Subject Sample	# 2	0.000	02:11
Breath Volume		2.750	02:11
Air Blank		0.000	02:12
Control Test		0.079	02:12
Air Blank		0.000	02:13
Diagnostics Check		OK	02:13

Data Download Date/Time: 01/12/2017 08:22

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROLLESTON,JACOB,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: ROLLESTON,JACOB
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 12/08/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: SNYDER

DATE/TIME OF TEST: 01/11/2017 21:23:33

Diagnostics Check		OK	21:20	
Air Blank		0.000	21:21	
Control Test		0.078	21:21	
Air Blank		0.000	21:21	
Subject Sample	# 1		21:22	Subject Test Refused
Breath Volume		0.000	21:22	
Air Blank		0.000	21:22	
Control Test		0.079	21:22	
Air Blank		0.000	21:23	
Diagnostics Check		OK	21:23	

Data Download Date/Time: 01/12/2017 08:22

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHEA,CPL,J
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: SHEA,JOHN
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 12/08/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:56

SUBJECT'S LAST NAME: REYNOSO-LUCIO

DATE/TIME OF TEST: 01/12/2017 02:24:54

Diagnostics Check		OK	02:18
Air Blank		0.000	02:19
Control Test		0.078	02:19
Air Blank		0.000	02:19
Subject Sample	# 1	0.136	02:20
Breath Volume		1.597	02:20
Air Blank		0.000	02:20
Air Blank		0.000	02:22
Subject Sample	# 2	0.144	02:23
Breath Volume		2.375	02:23
Air Blank		0.000	02:23
Control Test		0.078	02:24
Air Blank		0.000	02:24
Diagnostics Check		OK	02:24

Data Download Date/Time: 01/12/2017 08:22

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JASIUKIEWICZ,MICHAEL,P
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: BRUNER,SEAN
ARRESTING OFFICER AGENCY: Kenneth City PD
LAST AGENCY INSPECTION: 01/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: BALLIS

DATE/TIME OF TEST: 01/13/2017 04:09:06

Diagnostics Check		OK	04:05	
Air Blank		0.000	04:06	
Control Test		0.080	04:06	
Air Blank		0.000	04:07	
Subject Sample	# 1		04:07	Subject Test Refused
Breath Volume		0.000	04:07	
Air Blank		0.000	04:08	
Control Test		0.079	04:08	
Air Blank		0.000	04:09	
Diagnostics Check		OK	04:09	

Data Download Date/Time: 02/15/2017 08:46

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JASIUKIEWICZ,MICHAEL,P
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: JASIUKIEWICZ,MICHAEL
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 01/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:12

SUBJECT'S LAST NAME: MOORE

DATE/TIME OF TEST: 01/13/2017 05:39:52

Diagnostics Check		OK	05:32
Air Blank		0.000	05:33
Control Test		0.079	05:33
Air Blank		0.000	05:34
Subject Sample	# 1	0.198	05:35
Breath Volume		2.105	05:35
Air Blank		0.000	05:36
Air Blank		0.000	05:37
Subject Sample	# 2	0.198	05:38
Breath Volume		3.027	05:38
Air Blank		0.000	05:38
Control Test		0.078	05:39
Air Blank		0.000	05:39
Diagnostics Check		OK	05:39

Data Download Date/Time: 02/15/2017 08:46

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JASIUKIEWICZ,MICHAEL,P
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: JASIUKIEWICZ,MICHAEL
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 01/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:14

SUBJECT'S LAST NAME: SPEARS

DATE/TIME OF TEST: 01/14/2017 00:38:51

Diagnostics Check		OK	00:35	
Air Blank		0.000	00:36	
Control Test		0.079	00:36	
Air Blank		0.000	00:37	
Subject Sample	# 1		00:37	Subject Test Refused
Breath Volume		0.000	00:37	
Air Blank		0.000	00:37	
Control Test		0.079	00:38	
Air Blank		0.000	00:38	
Diagnostics Check		OK	00:38	

Data Download Date/Time: 02/15/2017 08:46

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROLLESTON,JACOB,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: ROLLESTON,JACOB
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 01/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: HARMON

DATE/TIME OF TEST: 01/14/2017 00:51:53

Diagnostics Check		OK	00:48	
Air Blank		0.000	00:49	
Control Test		0.080	00:49	
Air Blank		0.000	00:49	
Subject Sample	# 1		00:50	Subject Test Refused
Breath Volume		0.000	00:50	
Air Blank		0.000	00:50	
Control Test		0.079	00:51	
Air Blank		0.000	00:51	
Diagnostics Check		OK	00:51	

Data Download Date/Time: 02/15/2017 08:46

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JASIUKIEWICZ,MICHAEL,P
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: JASIUKIEWICZ,MICHAEL
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 01/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:47

SUBJECT'S LAST NAME: ELIAS

DATE/TIME OF TEST: 01/15/2017 02:14:08

Diagnostics Check		OK	02:07
Air Blank		0.000	02:08
Control Test		0.079	02:08
Air Blank		0.000	02:09
Subject Sample	# 1	0.166	02:09
Breath Volume		1.687	02:09
Air Blank		0.000	02:10
Air Blank		0.000	02:12
Subject Sample	# 2	0.161	02:12
Breath Volume		2.628	02:12
Air Blank		0.000	02:13
Control Test		0.079	02:13
Air Blank		0.000	02:14
Diagnostics Check		OK	02:14

Data Download Date/Time: 02/15/2017 08:46

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JASIUKIEWICZ,MICHAEL,P
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: JASIUKIEWICZ,MICHAEL
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 01/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:03

SUBJECT'S LAST NAME: NARVAEZ NARVAEZ

DATE/TIME OF TEST: 01/15/2017 03:30:19

Diagnostics Check		OK	03:23
Air Blank		0.000	03:24
Control Test		0.080	03:24
Air Blank		0.000	03:25
Subject Sample	# 1	0.158	03:25
Breath Volume		2.031	03:25
Air Blank		0.000	03:26
Air Blank		0.000	03:28
Subject Sample	# 2	0.161	03:28
Breath Volume		2.011	03:28
Air Blank		0.000	03:29
Control Test		0.078	03:29
Air Blank		0.000	03:30
Diagnostics Check		OK	03:30

Data Download Date/Time: 02/15/2017 08:46

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHEA,JOHN,D
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: SHEA,JOHN
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 01/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:33

SUBJECT'S LAST NAME: DICKINSON

DATE/TIME OF TEST: 01/15/2017 23:08:40

Diagnostics Check		OK	22:53	
Air Blank		0.000	22:54	
Control Test		0.078	22:54	
Air Blank		0.000	22:55	
Subject Sample	# 1	0.141	22:58	Volume Not Met
Breath Volume		0.605	22:58	
Air Blank		0.000	22:59	
Air Blank		0.000	23:00	
Subject Sample	# 2	0.174	23:02	
Breath Volume		2.281	23:02	
Air Blank		0.000	23:02	
Air Blank		0.000	23:04	
Subject Sample	# 3	0.165	23:06	
Breath Volume		1.839	23:06	
Air Blank		0.000	23:07	
Control Test		0.077	23:08	
Air Blank		0.000	23:08	
Diagnostics Check		OK	23:08	

Data Download Date/Time: 02/15/2017 08:46

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHEA,JOHN,D
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: SHEA,JOHN
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 01/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:48

SUBJECT'S LAST NAME: LEVY

DATE/TIME OF TEST: 01/16/2017 05:15:53

Diagnostics Check		OK	05:09
Air Blank		0.000	05:10
Control Test		0.079	05:10
Air Blank		0.000	05:10
Subject Sample	# 1	0.195	05:11
Breath Volume		3.050	05:11
Air Blank		0.000	05:12
Air Blank		0.000	05:13
Subject Sample	# 2	0.185	05:14
Breath Volume		2.187	05:14
Air Blank		0.000	05:14
Control Test		0.078	05:15
Air Blank		0.000	05:15
Diagnostics Check		OK	05:15

Data Download Date/Time: 02/15/2017 08:46

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACOBSEN,STEPHEN,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: JACOBSEN,STEPHEN
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 01/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:18

SUBJECT'S LAST NAME: SANTOS CASTRO

DATE/TIME OF TEST: 01/18/2017 23:44:28

Diagnostics Check		OK	23:38
Air Blank		0.000	23:38
Control Test		0.079	23:39
Air Blank		0.000	23:39
Subject Sample	# 1	0.000	23:40
Breath Volume		1.652	23:40
Air Blank		0.000	23:40
Air Blank		0.000	23:42
Subject Sample	# 2	0.000	23:43
Breath Volume		1.332	23:43
Air Blank		0.000	23:43
Control Test		0.079	23:43
Air Blank		0.000	23:44
Diagnostics Check		OK	23:44

Data Download Date/Time: 02/15/2017 08:46

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACOBSEN,STEPHEN,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: COPELAND,JEREMY
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 01/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:54

SUBJECT'S LAST NAME: YI

DATE/TIME OF TEST: 01/19/2017 02:29:42

Diagnostics Check		OK	02:22
Air Blank		0.000	02:22
Control Test		0.079	02:23
Air Blank		0.000	02:23
Subject Sample	# 1	0.131	02:25
Breath Volume		1.183	02:25
Air Blank		0.000	02:25
Air Blank		0.000	02:27
Subject Sample	# 2	0.149	02:28
Breath Volume		1.890	02:28
Air Blank		0.000	02:28
Control Test		0.077	02:29
Air Blank		0.000	02:29
Diagnostics Check		OK	02:29

Data Download Date/Time: 02/15/2017 08:46

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROLLESTON,JACOB,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: ROLLESTON,JACOB
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 01/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:20

SUBJECT'S LAST NAME: SIMONDS

DATE/TIME OF TEST: 01/19/2017 20:45:25

Diagnostics Check		OK	20:40	
Air Blank		0.000	20:41	
Control Test		0.079	20:41	
Air Blank		0.000	20:42	
Subject Sample	# 1		20:44	Subject Test Refused
Breath Volume		0.000	20:44	
Air Blank		0.000	20:44	
Control Test		0.079	20:44	
Air Blank		0.000	20:45	
Diagnostics Check		OK	20:45	

Data Download Date/Time: 02/15/2017 08:46

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROLLESTON,JACOB,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: ROLLESTON,JACOB
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 01/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:02

SUBJECT'S LAST NAME: WHIDDEN

DATE/TIME OF TEST: 01/21/2017 00:29:52

Diagnostics Check		OK	00:22
Air Blank		0.000	00:23
Control Test		0.079	00:23
Air Blank		0.000	00:24
Subject Sample	# 1	0.216	00:25
Breath Volume		1.117	00:25
Air Blank		0.000	00:25
Air Blank		0.000	00:27
Subject Sample	# 2	0.214	00:28
Breath Volume		2.042	00:28
Air Blank		0.000	00:28
Control Test		0.077	00:29
Air Blank		0.000	00:29
Diagnostics Check		OK	00:29

Data Download Date/Time: 02/15/2017 08:46

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JASIUKIEWICZ,MICHAEL,P
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: JASIUKIEWICZ,MICHAEL
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 01/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: GAROV

DATE/TIME OF TEST: 01/22/2017 02:28:49

Diagnostics Check		OK	02:22
Air Blank		0.000	02:22
Control Test		0.079	02:22
Air Blank		0.000	02:23
Subject Sample	# 1	0.153	02:24
Breath Volume		1.156	02:24
Air Blank		0.000	02:25
Air Blank		0.000	02:26
Subject Sample	# 2	0.149	02:27
Breath Volume		1.363	02:27
Air Blank		0.000	02:27
Control Test		0.078	02:28
Air Blank		0.000	02:28
Diagnostics Check		OK	02:28

Data Download Date/Time: 02/15/2017 08:46

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JASIUKIEWICZ,MICHAEL,P
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: BRUNER,S
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 01/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: SAENKA

DATE/TIME OF TEST: 01/22/2017 03:14:32

Diagnostics Check		OK	02:59	
Air Blank		0.000	03:00	
Control Test		0.078	03:00	
Air Blank		0.000	03:01	
Subject Sample	# 1	0.113	03:04	Volume Not Met
Breath Volume		0.828	03:04	
Air Blank		0.000	03:05	
Air Blank		0.000	03:06	
Subject Sample	# 2	0.125	03:07	
Breath Volume		1.445	03:07	
Air Blank		0.000	03:07	
Air Blank		0.000	03:09	
Subject Sample	# 3	0.114	03:12	Volume Not Met
Breath Volume		0.050	03:12	
Air Blank		0.000	03:13	
Control Test		0.077	03:13	
Air Blank		0.000	03:14	
Diagnostics Check		OK	03:14	

Data Download Date/Time: 02/15/2017 08:46

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOYLE,THOMAS,E
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: BOYLE,THOMAS
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 01/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:01

SUBJECT'S LAST NAME: HAYES

DATE/TIME OF TEST: 01/26/2017 23:28:22

Diagnostics Check		OK	23:22
Air Blank		0.000	23:22
Control Test		0.079	23:22
Air Blank		0.000	23:23
Subject Sample	# 1	0.000	23:24
Breath Volume		2.730	23:24
Air Blank		0.000	23:24
Air Blank		0.000	23:26
Subject Sample	# 2	0.000	23:27
Breath Volume		2.589	23:27
Air Blank		0.000	23:27
Control Test		0.080	23:27
Air Blank		0.000	23:28
Diagnostics Check		OK	23:28

Data Download Date/Time: 02/15/2017 08:46

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROLLESTON,JACOB,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: ROLLESTON,JACOB
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 01/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 17:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:04

SUBJECT'S LAST NAME: PAULSEN

DATE/TIME OF TEST: 01/27/2017 18:30:06

Diagnostics Check		OK	18:24
Air Blank		0.000	18:24
Control Test		0.080	18:24
Air Blank		0.000	18:25
Subject Sample	# 1	0.000	18:25
Breath Volume		3.339	18:25
Air Blank		0.000	18:26
Air Blank		0.000	18:28
Subject Sample	# 2	0.000	18:28
Breath Volume		2.984	18:28
Air Blank		0.000	18:29
Control Test		0.080	18:29
Air Blank		0.000	18:30
Diagnostics Check		OK	18:30

Data Download Date/Time: 02/15/2017 08:46

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JASIUKIEWICZ,MICHAEL,P
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: JASIUKIEWICZ,MICHAEL
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 01/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:59

SUBJECT'S LAST NAME: SHERRILL JR

DATE/TIME OF TEST: 01/27/2017 23:22:37

Diagnostics Check		OK	23:19	
Air Blank		0.000	23:20	
Control Test		0.080	23:20	
Air Blank		0.000	23:21	
Subject Sample	# 1		23:21	Subject Test Refused
Breath Volume		0.000	23:21	
Air Blank		0.000	23:21	
Control Test		0.080	23:22	
Air Blank		0.000	23:22	
Diagnostics Check		OK	23:22	

Data Download Date/Time: 02/15/2017 08:46

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JASIUKIEWICZ,MICHAEL,P
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: JASIUKIEWICZ,MICHAEL
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 01/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:51

SUBJECT'S LAST NAME: HAVERFIELD JR

DATE/TIME OF TEST: 01/29/2017 02:14:26

Diagnostics Check		OK	02:11	
Air Blank		0.000	02:12	
Control Test		0.080	02:12	
Air Blank		0.000	02:12	
Subject Sample	# 1		02:13	Subject Test Refused
Breath Volume		0.000	02:13	
Air Blank		0.000	02:13	
Control Test		0.079	02:13	
Air Blank		0.000	02:14	
Diagnostics Check		OK	02:14	

Data Download Date/Time: 02/15/2017 08:46

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROLLESTON,JACOB,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: ROLLESTON,JACOB
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 01/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:51

SUBJECT'S LAST NAME: MCCLELLAND

DATE/TIME OF TEST: 01/31/2017 22:18:07

Diagnostics Check		OK	22:11
Air Blank		0.000	22:12
Control Test		0.080	22:12
Air Blank		0.000	22:13
Subject Sample	# 1	0.317	22:13
Breath Volume		2.242	22:13
Air Blank		0.000	22:14
Air Blank		0.000	22:16
Subject Sample	# 2	0.316	22:16
Breath Volume		2.019	22:16
Air Blank		0.000	22:17
Control Test		0.078	22:17
Air Blank		0.000	22:18
Diagnostics Check		OK	22:18

Data Download Date/Time: 02/15/2017 08:46

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROLLESTON,JACOB,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: BRUNER,SEAN
ARRESTING OFFICER AGENCY: Kenneth City PD
LAST AGENCY INSPECTION: 01/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: FLAMURI

DATE/TIME OF TEST: 01/31/2017 23:38:05

Diagnostics Check		OK	23:35	
Air Blank		0.000	23:35	
Control Test		0.079	23:36	
Air Blank		0.000	23:36	
Subject Sample	# 1		23:36	Subject Test Refused
Breath Volume		0.000	23:36	
Air Blank		0.000	23:37	
Control Test		0.080	23:37	
Air Blank		0.000	23:38	
Diagnostics Check		OK	23:38	

Data Download Date/Time: 02/15/2017 08:46

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002097

INSTRUMENT REGISTERED TO: NAS JACKSONVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTHARD,CURTIS,J
BREATH TEST OPERATOR AGENCY: NAS Jacksonville PD
ARRESTING OFFICER: MCGINTY,MARK
ARRESTING OFFICER AGENCY: NAS Jacksonville PD
LAST AGENCY INSPECTION: 01/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13413080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:23

SUBJECT'S LAST NAME: SUTHARD

DATE/TIME OF TEST: 01/22/2017 21:58:34

Diagnostics Check		OK	21:50
Air Blank		0.000	21:50
Control Test		0.077	21:51
Air Blank		0.000	21:51
Subject Sample	# 1	0.230	21:52
Breath Volume		1.460	21:52
Air Blank		0.000	21:53
Air Blank		0.000	21:54
Subject Sample	# 2	0.210	21:56
Breath Volume		1.824	21:56
Air Blank		0.000	21:57
Control Test		0.078	21:57
Air Blank		0.000	21:58
Diagnostics Check		OK	21:58

Data Download Date/Time: 02/16/2017 11:22

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002232

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,JOSE,D
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: AIKEN,MATTHEW
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 12/20/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1826916
DRY GAS EXPIRY DATE: 03/09/2018
VIOLATION CODE: DUI
ARREST TIME: 06:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:20

SUBJECT'S LAST NAME: GARCIAOCAMPO

DATE/TIME OF TEST: 01/02/2017 07:49:08

Diagnostics Check		OK	07:46	
Air Blank		0.000	07:46	
Control Test		0.081	07:47	
Air Blank		0.000	07:47	
Subject Sample	# 1		07:47	Subject Test Refused
Breath Volume		0.000	07:47	
Air Blank		0.000	07:48	
Control Test		0.081	07:48	
Air Blank		0.000	07:49	
Diagnostics Check		OK	07:49	

Data Download Date/Time: 01/24/2017 10:51

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002232

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONGE,GRETTEL,M
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: MONGE,GRETTEL
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 12/20/2016

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1826916
DRY GAS EXPIRY DATE: 03/09/2018
VIOLATION CODE: DUI
ARREST TIME: 04:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:40

SUBJECT'S LAST NAME: CABRERA

DATE/TIME OF TEST: 01/05/2017 05:10:24

Diagnostics Check		OK	05:02
Air Blank		0.000	05:03
Control Test		0.080	05:03
Air Blank		0.000	05:03
Subject Sample	# 1	0.117	05:05
Breath Volume		1.968	05:05
Air Blank		0.000	05:06
Air Blank		0.000	05:07
Subject Sample	# 2	0.115	05:08
Breath Volume		1.738	05:08
Air Blank		0.000	05:09
Control Test		0.080	05:09
Air Blank		0.000	05:10
Diagnostics Check		OK	05:10

Data Download Date/Time: 01/24/2017 10:51

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002232

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,JOSE,D
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: LUKE,BRYANT
ARRESTING OFFICER AGENCY: Surfside PD
LAST AGENCY INSPECTION: 12/20/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1826916
DRY GAS EXPIRY DATE: 03/09/2018
VIOLATION CODE: DUI
ARREST TIME: 09:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:15

SUBJECT'S LAST NAME: MANN

DATE/TIME OF TEST: 01/07/2017 10:41:43

Diagnostics Check		OK	10:38	
Air Blank		0.000	10:39	
Control Test		0.080	10:39	
Air Blank		0.000	10:40	
Subject Sample	# 1		10:40	Subject Test Refused
Breath Volume		0.000	10:40	
Air Blank		0.000	10:40	
Control Test		0.080	10:41	
Air Blank		0.000	10:41	
Diagnostics Check		OK	10:41	

Data Download Date/Time: 01/24/2017 10:51

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002232

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CONFESSORE,STEVE,
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: RIESGO,LAZARO
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 01/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1826916
DRY GAS EXPIRY DATE: 03/09/2018
VIOLATION CODE: DUI
ARREST TIME: 02:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 01/27/2017 03:10:00

Diagnostics Check		OK	03:01	
Air Blank		0.000	03:01	
Control Test		0.075	03:02	
Air Blank		0.000	03:02	
Subject Sample	# 1	0.182	03:05	Volume Not Met
Breath Volume		0.089	03:05	
Air Blank		0.000	03:06	
Air Blank		0.000	03:08	
Subject Sample	# 2		03:08	Subject Test Refused
Breath Volume		0.000	03:08	
Air Blank		0.000	03:09	
Control Test		0.074	03:09	Control Outside Tolerance
Air Blank		0.000	03:10	

Data Download Date/Time: 02/28/2017 11:23

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002232

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CONFESSORE,STEVE,
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: COSNER,STEVE
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 01/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1826916
DRY GAS EXPIRY DATE: 03/09/2018
VIOLATION CODE: DUI
ARREST TIME: 04:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:25

SUBJECT'S LAST NAME: DELAHAYE

DATE/TIME OF TEST: 01/27/2017 04:58:09

Diagnostics Check		OK	04:49
Air Blank		0.000	04:50
Control Test		0.078	04:50
Air Blank		0.000	04:51
Subject Sample	# 1	0.180	04:52
Breath Volume		1.984	04:52
Air Blank		0.000	04:53
Air Blank		0.000	04:54
Subject Sample	# 2	0.173	04:56
Breath Volume		2.535	04:56
Air Blank		0.000	04:57
Control Test		0.076	04:57
Air Blank		0.000	04:58
Diagnostics Check		OK	04:58

Data Download Date/Time: 02/28/2017 11:23

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002232

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERNANDEZ,JAVIER,
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: GONZALEZ,BRYAN
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 01/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1826916
DRY GAS EXPIRY DATE: 03/09/2018
VIOLATION CODE: DUI
ARREST TIME: 18:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:25

SUBJECT'S LAST NAME: AMIN AGLAN

DATE/TIME OF TEST: 01/29/2017 20:58:10

Diagnostics Check		OK	20:51
Air Blank		0.000	20:52
Control Test		0.080	20:52
Air Blank		0.000	20:52
Subject Sample	# 1	0.255	20:53
Breath Volume		2.386	20:53
Air Blank		0.000	20:54
Air Blank		0.000	20:56
Subject Sample	# 2	0.252	20:56
Breath Volume		2.394	20:56
Air Blank		0.000	20:57
Control Test		0.079	20:57
Air Blank		0.000	20:58
Diagnostics Check		OK	20:58

Data Download Date/Time: 02/28/2017 11:23

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002232

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CREWS,CURTIS,L
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: RODRIGUEZ,HUMBERTO
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 01/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1826916
DRY GAS EXPIRY DATE: 03/09/2018
VIOLATION CODE: DUI
ARREST TIME: 02:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: FILIUS

DATE/TIME OF TEST: 01/30/2017 03:19:47

Diagnostics Check		OK	03:16	
Air Blank		0.000	03:16	
Control Test		0.080	03:16	
Air Blank		0.000	03:17	
Subject Sample	# 1		03:18	Subject Test Refused
Breath Volume		0.000	03:18	
Air Blank		0.000	03:18	
Control Test		0.081	03:19	
Air Blank		0.000	03:19	
Diagnostics Check		OK	03:19	

Data Download Date/Time: 02/28/2017 11:23

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002463

INSTRUMENT REGISTERED TO: HIALEAH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARROJAS,GABRIEL,L
BREATH TEST OPERATOR AGENCY: Hialeah PD
ARRESTING OFFICER: ELIAS,FELIX
ARRESTING OFFICER AGENCY: Hialeah PD
LAST AGENCY INSPECTION: 01/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06415080A1
DRY GAS EXPIRY DATE: 04/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 01/29/2017 02:26:34

Diagnostics Check		OK	02:22	
Air Blank		0.000	02:23	
Control Test		0.077	02:23	
Air Blank		0.000	02:23	
Subject Sample	# 1		02:25	Subject Test Refused
Breath Volume		0.000	02:25	
Air Blank		0.000	02:25	
Control Test		0.078	02:25	
Air Blank		0.000	02:26	
Diagnostics Check		OK	02:26	

Data Download Date/Time: 02/23/2017 12:33

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002989

INSTRUMENT REGISTERED TO: TALLAHASSEE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KELLER,DAVID,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: SODERHOLM,CHELSEA
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 12/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: JENKINS

DATE/TIME OF TEST: 01/06/2017 04:09:41

Diagnostics Check		OK	04:03
Air Blank		0.000	04:03
Control Test		0.079	04:04
Air Blank		0.000	04:04
Subject Sample	# 1	0.135	04:05
Breath Volume		1.765	04:05
Air Blank		0.000	04:05
Air Blank		0.000	04:07
Subject Sample	# 2	0.136	04:08
Breath Volume		2.246	04:08
Air Blank		0.000	04:08
Control Test		0.078	04:09
Air Blank		0.000	04:09
Diagnostics Check		OK	04:09

Data Download Date/Time: 01/31/2017 14:57

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 19:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:19

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 01/10/2017 20:47:09

Diagnostics Check		OK	20:39
Air Blank		0.000	20:40
Control Test		0.079	20:40
Air Blank		0.000	20:41
Subject Sample	# 1	0.247	20:41
Breath Volume		1.878	20:41
Air Blank		0.000	20:42
Air Blank		0.000	20:44
Subject Sample	# 2	0.245	20:45
Breath Volume		2.285	20:45
Air Blank		0.000	20:46
Control Test		0.078	20:46
Air Blank		0.000	20:47
Diagnostics Check		OK	20:47

Data Download Date/Time: 02/08/2017 19:30

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LARA,A
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: HEGGER

DATE/TIME OF TEST: 01/10/2017 23:27:34

Diagnostics Check		OK	23:20
Air Blank		0.000	23:21
Control Test		0.079	23:21
Air Blank		0.000	23:22
Subject Sample	# 1	0.153	23:22
Breath Volume		1.324	23:22
Air Blank		0.000	23:23
Air Blank		0.000	23:24
Subject Sample	# 2	0.162	23:25
Breath Volume		1.316	23:25
Air Blank		0.000	23:26
Control Test		0.079	23:26
Air Blank		0.000	23:27
Diagnostics Check		OK	23:27

Data Download Date/Time: 02/08/2017 19:30

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:09

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 01/11/2017 04:37:27

Diagnostics Check		OK	04:29
Air Blank		0.000	04:29
Control Test		0.079	04:30
Air Blank		0.000	04:30
Subject Sample	# 1	0.165	04:31
Breath Volume		1.574	04:31
Air Blank		0.000	04:31
Air Blank		0.000	04:33
Subject Sample	# 2	0.161	04:35
Breath Volume		1.328	04:35
Air Blank		0.000	04:36
Control Test		0.079	04:36
Air Blank		0.000	04:37
Diagnostics Check		OK	04:37

Data Download Date/Time: 02/08/2017 19:30

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:22

SUBJECT'S LAST NAME: TERRIBLES

DATE/TIME OF TEST: 01/11/2017 21:48:46

Diagnostics Check		OK	21:42
Air Blank		0.000	21:43
Control Test		0.078	21:43
Air Blank		0.000	21:43
Subject Sample	# 1	0.159	21:44
Breath Volume		2.531	21:44
Air Blank		0.000	21:44
Air Blank		0.000	21:46
Subject Sample	# 2	0.163	21:47
Breath Volume		1.484	21:47
Air Blank		0.000	21:47
Control Test		0.078	21:48
Air Blank		0.000	21:48
Diagnostics Check		OK	21:48

Data Download Date/Time: 02/08/2017 19:30

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: NIEUWENDAM,R
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:23

SUBJECT'S LAST NAME: GREAR

DATE/TIME OF TEST: 01/11/2017 23:50:13

Diagnostics Check		OK	23:43
Air Blank		0.000	23:44
Control Test		0.079	23:44
Air Blank		0.000	23:45
Subject Sample	# 1	0.100	23:45
Breath Volume		2.089	23:45
Air Blank		0.000	23:46
Air Blank		0.000	23:48
Subject Sample	# 2	0.097	23:48
Breath Volume		1.859	23:48
Air Blank		0.000	23:49
Control Test		0.078	23:49
Air Blank		0.000	23:50
Diagnostics Check		OK	23:50

Data Download Date/Time: 02/08/2017 19:30

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY,PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DEJESUS,V
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:01

SUBJECT'S LAST NAME: JUNIOR

DATE/TIME OF TEST: 01/12/2017 22:29:47

Diagnostics Check		OK	22:22
Air Blank		0.000	22:23
Control Test		0.079	22:23
Air Blank		0.000	22:24
Subject Sample	# 1	0.121	22:25
Breath Volume		1.597	22:25
Air Blank		0.000	22:25
Air Blank		0.000	22:27
Subject Sample	# 2	0.116	22:28
Breath Volume		1.468	22:28
Air Blank		0.000	22:28
Control Test		0.078	22:29
Air Blank		0.000	22:29
Diagnostics Check		OK	22:29

Data Download Date/Time: 02/08/2017 19:30

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FRANKLIN,R
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:33

SUBJECT'S LAST NAME: RIVERA MEDINA

DATE/TIME OF TEST: 01/13/2017 02:59:58

Diagnostics Check		OK	02:53	
Air Blank		0.000	02:54	
Control Test		0.079	02:54	
Air Blank		0.000	02:54	
Subject Sample	# 1	0.177	02:55	
Breath Volume		1.699	02:55	
Air Blank		0.000	02:56	
Air Blank		0.000	02:57	
Subject Sample	# 2		02:58	Slope Not Met
Breath Volume		1.121	02:58	
Air Blank		0.000	02:59	
Control Test		0.078	02:59	
Air Blank		0.000	02:59	
Diagnostics Check		OK	02:59	

Data Download Date/Time: 02/08/2017 19:30

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FRANKLIN,R
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:58

SUBJECT'S LAST NAME: RIVERA MEDINA

DATE/TIME OF TEST: 01/13/2017 03:25:48

Diagnostics Check		OK	03:18
Air Blank		0.000	03:19
Control Test		0.078	03:19
Air Blank		0.000	03:19
Subject Sample	# 1	0.166	03:20
Breath Volume		1.628	03:20
Air Blank		0.000	03:21
Air Blank		0.000	03:23
Subject Sample	# 2	0.165	03:24
Breath Volume		2.394	03:24
Air Blank		0.000	03:24
Control Test		0.078	03:25
Air Blank		0.000	03:25
Diagnostics Check		OK	03:25

Data Download Date/Time: 02/08/2017 19:30

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRABAND,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:26

SUBJECT'S LAST NAME: ALANIS

DATE/TIME OF TEST: 01/13/2017 03:54:31

Diagnostics Check		OK	03:47
Air Blank		0.000	03:48
Control Test		0.078	03:48
Air Blank		0.000	03:48
Subject Sample	# 1	0.184	03:49
Breath Volume		1.273	03:49
Air Blank		0.000	03:50
Air Blank		0.000	03:52
Subject Sample	# 2	0.172	03:52
Breath Volume		1.371	03:52
Air Blank		0.000	03:53
Control Test		0.078	03:53
Air Blank		0.000	03:54
Diagnostics Check		OK	03:54

Data Download Date/Time: 02/08/2017 19:30

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRIMBLE,TERESA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VAN-TREESE,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:44

SUBJECT'S LAST NAME: LUBIN

DATE/TIME OF TEST: 01/13/2017 23:11:49

Diagnostics Check		OK	23:04
Air Blank		0.000	23:05
Control Test		0.079	23:05
Air Blank		0.000	23:05
Subject Sample	# 1	0.108	23:07
Breath Volume		1.886	23:07
Air Blank		0.000	23:07
Air Blank		0.000	23:09
Subject Sample	# 2	0.113	23:10
Breath Volume		1.601	23:10
Air Blank		0.000	23:10
Control Test		0.078	23:11
Air Blank		0.000	23:11
Diagnostics Check		OK	23:11

Data Download Date/Time: 02/08/2017 19:30

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY,PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SIMPKINS,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:16

SUBJECT'S LAST NAME: GRACIA PETTERSON

DATE/TIME OF TEST: 01/14/2017 00:44:11

Diagnostics Check		OK	00:36
Air Blank		0.000	00:37
Control Test		0.079	00:37
Air Blank		0.000	00:37
Subject Sample	# 1	0.155	00:38
Breath Volume		1.343	00:38
Air Blank		0.000	00:38
Air Blank		0.000	00:40
Subject Sample	# 2	0.153	00:42
Breath Volume		1.910	00:42
Air Blank		0.000	00:43
Control Test		0.079	00:43
Air Blank		0.000	00:44
Diagnostics Check		OK	00:44

Data Download Date/Time: 02/08/2017 19:30

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: NIEUWENDAM,R
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:44

SUBJECT'S LAST NAME: COSTANZA

DATE/TIME OF TEST: 01/14/2017 04:11:02

Diagnostics Check		OK	04:04
Air Blank		0.000	04:05
Control Test		0.078	04:05
Air Blank		0.000	04:06
Subject Sample	# 1	0.151	04:06
Breath Volume		1.886	04:06
Air Blank		0.000	04:07
Air Blank		0.000	04:09
Subject Sample	# 2	0.151	04:09
Breath Volume		2.472	04:09
Air Blank		0.000	04:10
Control Test		0.078	04:10
Air Blank		0.000	04:10
Diagnostics Check		OK	04:11

Data Download Date/Time: 02/08/2017 19:30

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: EDGE,S
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 19:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:35

SUBJECT'S LAST NAME: LUIS-JUAN

DATE/TIME OF TEST: 01/14/2017 21:02:17

Diagnostics Check		OK	20:56
Air Blank		0.000	20:56
Control Test		0.079	20:56
Air Blank		0.000	20:57
Subject Sample	# 1	0.243	20:57
Breath Volume		1.144	20:57
Air Blank		0.000	20:58
Air Blank		0.000	21:00
Subject Sample	# 2	0.239	21:00
Breath Volume		1.582	21:00
Air Blank		0.000	21:01
Control Test		0.077	21:01
Air Blank		0.000	21:02
Diagnostics Check		OK	21:02

Data Download Date/Time: 02/08/2017 19:30

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PECYLAK,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:37

SUBJECT'S LAST NAME: BRUNO

DATE/TIME OF TEST: 01/15/2017 00:06:23

Diagnostics Check		OK	23:57	
Air Blank		0.000	23:58	
Control Test		0.079	23:58	
Air Blank		0.000	23:59	
Subject Sample	# 1		00:02	No Sample Provided
Breath Volume		0.000	00:02	
Air Blank		0.000	00:02	
Air Blank		0.000	00:04	
Subject Sample	# 2		00:05	Subject Test Refused
Breath Volume		0.000	00:05	
Air Blank		0.000	00:05	
Control Test		0.078	00:05	
Air Blank		0.000	00:06	
Diagnostics Check		OK	00:06	

Data Download Date/Time: 02/08/2017 19:30

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SIMPKINS,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:48

SUBJECT'S LAST NAME: REED

DATE/TIME OF TEST: 01/15/2017 03:17:19

Diagnostics Check		OK	03:10
Air Blank		0.000	03:11
Control Test		0.079	03:11
Air Blank		0.000	03:12
Subject Sample	# 1	0.128	03:12
Breath Volume		4.023	03:12
Air Blank		0.000	03:13
Air Blank		0.000	03:15
Subject Sample	# 2	0.128	03:15
Breath Volume		3.792	03:15
Air Blank		0.000	03:16
Control Test		0.078	03:16
Air Blank		0.000	03:17
Diagnostics Check		OK	03:17

Data Download Date/Time: 02/08/2017 19:30

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:52

SUBJECT'S LAST NAME: SOTO-MARIN

DATE/TIME OF TEST: 01/15/2017 04:19:59

Diagnostics Check		OK	04:12
Air Blank		0.000	04:13
Control Test		0.079	04:13
Air Blank		0.000	04:14
Subject Sample	# 1	0.113	04:15
Breath Volume		1.589	04:15
Air Blank		0.000	04:15
Air Blank		0.000	04:17
Subject Sample	# 2	0.111	04:18
Breath Volume		2.179	04:18
Air Blank		0.000	04:19
Control Test		0.078	04:19
Air Blank		0.000	04:19
Diagnostics Check		OK	04:19

Data Download Date/Time: 02/08/2017 19:30

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VELAZQUEZ,B
ARRESTING OFFICER AGENCY: U of So FL Campus PD
LAST AGENCY INSPECTION: 01/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:16

SUBJECT'S LAST NAME: CHAKRABARTI

DATE/TIME OF TEST: 01/15/2017 04:44:46

Diagnostics Check		OK	04:38
Air Blank		0.000	04:38
Control Test		0.078	04:39
Air Blank		0.000	04:39
Subject Sample	# 1	0.109	04:40
Breath Volume		3.039	04:40
Air Blank		0.000	04:40
Air Blank		0.000	04:42
Subject Sample	# 2	0.111	04:43
Breath Volume		2.769	04:43
Air Blank		0.000	04:43
Control Test		0.078	04:44
Air Blank		0.000	04:44
Diagnostics Check		OK	04:44

Data Download Date/Time: 02/08/2017 19:30

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DUNCAN,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 18:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:58

SUBJECT'S LAST NAME: LEE

DATE/TIME OF TEST: 01/15/2017 19:25:52

Diagnostics Check		OK	19:18
Air Blank		0.000	19:19
Control Test		0.079	19:19
Air Blank		0.000	19:20
Subject Sample	# 1	0.188	19:20
Breath Volume		1.328	19:20
Air Blank		0.000	19:21
Air Blank		0.000	19:23
Subject Sample	# 2	0.197	19:24
Breath Volume		1.785	19:24
Air Blank		0.000	19:24
Control Test		0.079	19:25
Air Blank		0.000	19:25
Diagnostics Check		OK	19:25

Data Download Date/Time: 02/08/2017 19:30

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:12

SUBJECT'S LAST NAME: MEJIA-ROSADO

DATE/TIME OF TEST: 01/16/2017 02:35:10

Diagnostics Check		OK	02:32	
Air Blank		0.000	02:32	
Control Test		0.079	02:33	
Air Blank		0.000	02:33	
Subject Sample	# 1		02:33	Subject Test Refused
Breath Volume		0.000	02:33	
Air Blank		0.000	02:34	
Control Test		0.079	02:34	
Air Blank		0.000	02:35	
Diagnostics Check		OK	02:35	

Data Download Date/Time: 02/08/2017 19:30

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNYDER,MELONIE,E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: EHRHART,M
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 16:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:12

SUBJECT'S LAST NAME: PAPA

DATE/TIME OF TEST: 01/16/2017 17:40:21

Diagnostics Check		OK	17:33
Air Blank		0.000	17:34
Control Test		0.079	17:34
Air Blank		0.000	17:35
Subject Sample	# 1	0.336	17:35
Breath Volume		2.433	17:35
Air Blank		0.000	17:36
Air Blank		0.000	17:38
Subject Sample	# 2	0.335	17:38
Breath Volume		2.218	17:38
Air Blank		0.000	17:39
Control Test		0.078	17:39
Air Blank		0.000	17:40
Diagnostics Check		OK	17:40

Data Download Date/Time: 02/08/2017 19:30

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GONZALEZ,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 19:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:34

SUBJECT'S LAST NAME: VICTOR

DATE/TIME OF TEST: 01/17/2017 20:03:09

Diagnostics Check		OK	19:56
Air Blank		0.000	19:57
Control Test		0.079	19:57
Air Blank		0.000	19:57
Subject Sample	# 1	0.116	19:58
Breath Volume		1.816	19:58
Air Blank		0.000	19:58
Air Blank		0.000	20:00
Subject Sample	# 2	0.114	20:01
Breath Volume		2.777	20:01
Air Blank		0.000	20:02
Control Test		0.079	20:02
Air Blank		0.000	20:03
Diagnostics Check		OK	20:03

Data Download Date/Time: 02/08/2017 19:30

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRIMBLE,TERESA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: NIEUWENDAM,R
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: STROMAN

DATE/TIME OF TEST: 01/18/2017 02:21:04

Diagnostics Check		OK	02:11	
Air Blank		0.000	02:11	
Control Test		0.078	02:11	
Air Blank		0.000	02:12	
Subject Sample	# 1	0.122	02:15	Volume Not Met
Breath Volume		0.300	02:15	
Air Blank		0.000	02:16	
Air Blank		0.000	02:17	
Subject Sample	# 2		02:19	Subject Test Refused
Breath Volume		0.066	02:19	
Air Blank		0.000	02:20	
Control Test		0.078	02:20	
Air Blank		0.000	02:20	
Diagnostics Check		OK	02:21	

Data Download Date/Time: 02/08/2017 19:30

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARR,BARBARA,H
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: UNO,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 14:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:03

SUBJECT'S LAST NAME: MISCHO

DATE/TIME OF TEST: 01/18/2017 15:30:21

Diagnostics Check		OK	15:24
Air Blank		0.000	15:25
Control Test		0.079	15:25
Air Blank		0.000	15:25
Subject Sample	# 1	0.000	15:26
Breath Volume		2.304	15:26
Air Blank		0.000	15:26
Air Blank		0.000	15:28
Subject Sample	# 2	0.000	15:28
Breath Volume		2.996	15:28
Air Blank		0.000	15:29
Control Test		0.078	15:29
Air Blank		0.000	15:30
Diagnostics Check		OK	15:30

Data Download Date/Time: 02/08/2017 19:30

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY,PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:53

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 01/18/2017 21:19:57

Diagnostics Check		OK	21:13
Air Blank		0.000	21:14
Control Test		0.079	21:14
Air Blank		0.000	21:14
Subject Sample	# 1	0.247	21:15
Breath Volume		1.808	21:15
Air Blank		0.000	21:16
Air Blank		0.000	21:17
Subject Sample	# 2	0.249	21:18
Breath Volume		2.035	21:18
Air Blank		0.000	21:19
Control Test		0.078	21:19
Air Blank		0.000	21:19
Diagnostics Check		OK	21:19

Data Download Date/Time: 02/08/2017 19:30

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PECYLAK,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:56

SUBJECT'S LAST NAME: HUDSON

DATE/TIME OF TEST: 01/20/2017 03:28:07

Diagnostics Check		OK	03:16	
Air Blank		0.000	03:16	
Control Test		0.080	03:17	
Air Blank		0.000	03:17	
Subject Sample	# 1	0.125	03:20	Volume Not Met
Breath Volume		0.070	03:20	
Air Blank		0.000	03:21	
Air Blank		0.000	03:23	
Subject Sample	# 2	0.100	03:26	Volume Not Met
Breath Volume		0.082	03:26	
Air Blank		0.000	03:27	
Control Test		0.078	03:27	
Air Blank		0.000	03:28	
Diagnostics Check		OK	03:28	

Data Download Date/Time: 02/08/2017 19:31

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PECYLAK,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:56

SUBJECT'S LAST NAME: HUDSON

DATE/TIME OF TEST: 01/20/2017 03:43:05

Diagnostics Check		OK	03:31	
Air Blank		0.000	03:31	
Control Test		0.078	03:32	
Air Blank		0.000	03:32	
Subject Sample	# 1	0.096	03:35	Volume Not Met
Breath Volume		0.050	03:35	
Air Blank		0.000	03:36	
Air Blank		0.000	03:38	
Subject Sample	# 2	0.093	03:41	Volume Not Met
Breath Volume		0.210	03:41	
Air Blank		0.000	03:42	
Control Test		0.077	03:42	
Air Blank		0.000	03:43	
Diagnostics Check		OK	03:43	

Data Download Date/Time: 02/08/2017 19:31

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:41

SUBJECT'S LAST NAME: HAKIM

DATE/TIME OF TEST: 01/20/2017 04:09:38

Diagnostics Check		OK	04:02
Air Blank		0.000	04:02
Control Test		0.078	04:03
Air Blank		0.000	04:03
Subject Sample	# 1	0.147	04:04
Breath Volume		1.261	04:04
Air Blank		0.000	04:05
Air Blank		0.000	04:06
Subject Sample	# 2	0.154	04:08
Breath Volume		1.257	04:08
Air Blank		0.000	04:08
Control Test		0.078	04:09
Air Blank		0.000	04:09
Diagnostics Check		OK	04:09

Data Download Date/Time: 02/08/2017 19:31

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRIMBLE,TERESA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:24

SUBJECT'S LAST NAME: MACFARLANE

DATE/TIME OF TEST: 01/20/2017 21:49:33

Diagnostics Check		OK	21:44	
Air Blank		0.000	21:45	
Control Test		0.079	21:45	
Air Blank		0.000	21:45	
Subject Sample	# 1		21:47	Subject Test Refused
Breath Volume		0.000	21:47	
Air Blank		0.000	21:48	
Control Test		0.079	21:48	
Air Blank		0.000	21:49	
Diagnostics Check		OK	21:49	

Data Download Date/Time: 02/08/2017 19:31

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: TAVERAS,N
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:17

SUBJECT'S LAST NAME: ROMERO-MALDONADO

DATE/TIME OF TEST: 01/20/2017 23:45:01

Diagnostics Check		OK	23:38
Air Blank		0.000	23:39
Control Test		0.079	23:39
Air Blank		0.000	23:40
Subject Sample	# 1	0.160	23:40
Breath Volume		1.277	23:40
Air Blank		0.000	23:41
Air Blank		0.000	23:42
Subject Sample	# 2	0.168	23:43
Breath Volume		1.445	23:43
Air Blank		0.000	23:44
Control Test		0.079	23:44
Air Blank		0.000	23:44
Diagnostics Check		OK	23:45

Data Download Date/Time: 02/08/2017 19:31

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LYON,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:16

SUBJECT'S LAST NAME: CALLE

DATE/TIME OF TEST: 01/21/2017 04:42:54

Diagnostics Check		OK	04:36
Air Blank		0.000	04:36
Control Test		0.078	04:36
Air Blank		0.000	04:37
Subject Sample	# 1	0.141	04:38
Breath Volume		1.390	04:38
Air Blank		0.000	04:38
Air Blank		0.000	04:40
Subject Sample	# 2	0.143	04:41
Breath Volume		1.429	04:41
Air Blank		0.000	04:41
Control Test		0.078	04:42
Air Blank		0.000	04:42
Diagnostics Check		OK	04:42

Data Download Date/Time: 02/08/2017 19:31

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY,PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRENNAN,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:22

SUBJECT'S LAST NAME: VILLEGAS-MEDRANO

DATE/TIME OF TEST: 01/21/2017 22:49:34

Diagnostics Check		OK	22:42
Air Blank		0.000	22:43
Control Test		0.079	22:43
Air Blank		0.000	22:44
Subject Sample	# 1	0.178	22:45
Breath Volume		2.097	22:45
Air Blank		0.000	22:45
Air Blank		0.000	22:47
Subject Sample	# 2	0.182	22:48
Breath Volume		1.750	22:48
Air Blank		0.000	22:48
Control Test		0.077	22:48
Air Blank		0.000	22:49
Diagnostics Check		OK	22:49

Data Download Date/Time: 02/08/2017 19:31

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRIMBLE,TERESA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: CRAMER,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:01

SUBJECT'S LAST NAME: FLORES PACHECO

DATE/TIME OF TEST: 01/22/2017 01:24:27

Diagnostics Check		OK	01:21	
Air Blank		0.000	01:22	
Control Test		0.079	01:22	
Air Blank		0.000	01:22	
Subject Sample	# 1		01:23	Subject Test Refused
Breath Volume		0.000	01:23	
Air Blank		0.000	01:23	
Control Test		0.079	01:23	
Air Blank		0.000	01:24	
Diagnostics Check		OK	01:24	

Data Download Date/Time: 02/08/2017 19:31

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY,PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRABAND,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:48

SUBJECT'S LAST NAME: SCOTT SR

DATE/TIME OF TEST: 01/22/2017 03:15:52

Diagnostics Check		OK	03:08
Air Blank		0.000	03:08
Control Test		0.079	03:09
Air Blank		0.000	03:09
Subject Sample	# 1	0.118	03:11
Breath Volume		1.324	03:11
Air Blank		0.000	03:11
Air Blank		0.000	03:13
Subject Sample	# 2	0.119	03:14
Breath Volume		1.261	03:14
Air Blank		0.000	03:14
Control Test		0.077	03:15
Air Blank		0.000	03:15
Diagnostics Check		OK	03:15

Data Download Date/Time: 02/08/2017 19:31

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRIMBLE,TERESA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DOKENDORF,B
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: FERGUSON

DATE/TIME OF TEST: 01/22/2017 03:48:43

Diagnostics Check		OK	03:42
Air Blank		0.000	03:42
Control Test		0.078	03:43
Air Blank		0.000	03:43
Subject Sample	# 1	0.140	03:44
Breath Volume		1.234	03:44
Air Blank		0.000	03:44
Air Blank		0.000	03:46
Subject Sample	# 2	0.138	03:47
Breath Volume		1.550	03:47
Air Blank		0.000	03:47
Control Test		0.078	03:48
Air Blank		0.000	03:48
Diagnostics Check		OK	03:48

Data Download Date/Time: 02/08/2017 19:31

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRENNAN,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:23

SUBJECT'S LAST NAME: PINERO

DATE/TIME OF TEST: 01/22/2017 04:50:22

Diagnostics Check		OK	04:43
Air Blank		0.000	04:44
Control Test		0.079	04:44
Air Blank		0.000	04:44
Subject Sample	# 1	0.211	04:45
Breath Volume		2.496	04:45
Air Blank		0.000	04:46
Air Blank		0.000	04:48
Subject Sample	# 2	0.213	04:48
Breath Volume		2.433	04:48
Air Blank		0.000	04:49
Control Test		0.078	04:49
Air Blank		0.000	04:50
Diagnostics Check		OK	04:50

Data Download Date/Time: 02/08/2017 19:31

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY,PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRABAND,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 05:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:48

SUBJECT'S LAST NAME: NAJAR

DATE/TIME OF TEST: 01/22/2017 06:14:54

Diagnostics Check		OK	06:08
Air Blank		0.000	06:08
Control Test		0.079	06:08
Air Blank		0.000	06:09
Subject Sample	# 1	0.168	06:10
Breath Volume		1.394	06:10
Air Blank		0.000	06:11
Air Blank		0.000	06:12
Subject Sample	# 2	0.175	06:13
Breath Volume		1.703	06:13
Air Blank		0.000	06:13
Control Test		0.077	06:14
Air Blank		0.000	06:14
Diagnostics Check		OK	06:14

Data Download Date/Time: 02/08/2017 19:31

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRIMBLE,TERESA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRENNAN,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 18:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:50

SUBJECT'S LAST NAME: TORRES

DATE/TIME OF TEST: 01/22/2017 19:19:07

Diagnostics Check		OK	19:11
Air Blank		0.000	19:12
Control Test		0.078	19:12
Air Blank		0.000	19:12
Subject Sample	# 1	0.247	19:14
Breath Volume		1.171	19:14
Air Blank		0.000	19:14
Air Blank		0.000	19:16
Subject Sample	# 2	0.252	19:17
Breath Volume		2.554	19:17
Air Blank		0.000	19:18
Control Test		0.077	19:18
Air Blank		0.000	19:19
Diagnostics Check		OK	19:19

Data Download Date/Time: 02/08/2017 19:31

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRIMBLE,TERESA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARFIELD CRAIG,C
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:33

SUBJECT'S LAST NAME: ORTEGA

DATE/TIME OF TEST: 01/24/2017 05:00:00

Diagnostics Check		OK	04:53
Air Blank		0.000	04:53
Control Test		0.079	04:54
Air Blank		0.000	04:54
Subject Sample	# 1	0.123	04:55
Breath Volume		1.289	04:55
Air Blank		0.000	04:56
Air Blank		0.000	04:57
Subject Sample	# 2	0.125	04:58
Breath Volume		2.265	04:58
Air Blank		0.000	04:59
Control Test		0.078	04:59
Air Blank		0.000	04:59
Diagnostics Check		OK	05:00

Data Download Date/Time: 02/08/2017 19:31

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:22

SUBJECT'S LAST NAME: MATSON

DATE/TIME OF TEST: 01/25/2017 03:48:39

Diagnostics Check		OK	03:42
Air Blank		0.000	03:43
Control Test		0.080	03:43
Air Blank		0.000	03:43
Subject Sample	# 1	0.075	03:44
Breath Volume		1.179	03:44
Air Blank		0.000	03:44
Air Blank		0.000	03:46
Subject Sample	# 2	0.071	03:47
Breath Volume		1.277	03:47
Air Blank		0.000	03:47
Control Test		0.080	03:48
Air Blank		0.000	03:48
Diagnostics Check		OK	03:48

Data Download Date/Time: 02/08/2017 19:31

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GONZALEZ,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:56

SUBJECT'S LAST NAME: BRYANT

DATE/TIME OF TEST: 01/26/2017 02:23:10

Diagnostics Check		OK	02:16
Air Blank		0.000	02:17
Control Test		0.079	02:17
Air Blank		0.000	02:18
Subject Sample	# 1	0.138	02:18
Breath Volume		2.113	02:18
Air Blank		0.000	02:19
Air Blank		0.000	02:21
Subject Sample	# 2	0.135	02:21
Breath Volume		1.195	02:21
Air Blank		0.000	02:22
Control Test		0.079	02:22
Air Blank		0.000	02:23
Diagnostics Check		OK	02:23

Data Download Date/Time: 02/08/2017 19:31

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRENNAN,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:45

SUBJECT'S LAST NAME: DE LA CRUZ

DATE/TIME OF TEST: 01/26/2017 05:13:11

Diagnostics Check		OK	05:06
Air Blank		0.000	05:06
Control Test		0.080	05:06
Air Blank		0.000	05:07
Subject Sample	# 1	0.190	05:08
Breath Volume		2.519	05:08
Air Blank		0.000	05:09
Air Blank		0.000	05:10
Subject Sample	# 2	0.192	05:11
Breath Volume		2.445	05:11
Air Blank		0.000	05:12
Control Test		0.078	05:12
Air Blank		0.000	05:13
Diagnostics Check		OK	05:13

Data Download Date/Time: 02/08/2017 19:31

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MONTOUTE,S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:01

SUBJECT'S LAST NAME: VELEZ

DATE/TIME OF TEST: 01/27/2017 00:27:53

Diagnostics Check		OK	00:21
Air Blank		0.000	00:22
Control Test		0.078	00:22
Air Blank		0.000	00:22
Subject Sample	# 1	0.179	00:23
Breath Volume		1.917	00:23
Air Blank		0.000	00:24
Air Blank		0.000	00:25
Subject Sample	# 2	0.172	00:26
Breath Volume		1.835	00:26
Air Blank		0.000	00:26
Control Test		0.079	00:27
Air Blank		0.000	00:27
Diagnostics Check		OK	00:27

Data Download Date/Time: 02/08/2017 19:31

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRIMBLE,TERESA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:44

SUBJECT'S LAST NAME: GREIG

DATE/TIME OF TEST: 01/27/2017 01:25:25

Diagnostics Check		OK	01:13	
Air Blank		0.000	01:14	
Control Test		0.078	01:14	
Air Blank		0.000	01:14	
Subject Sample	# 1	0.165	01:18	Volume Not Met
Breath Volume		0.660	01:18	
Air Blank		0.000	01:18	
Air Blank		0.000	01:20	
Subject Sample	# 2	0.156	01:23	Volume Not Met
Breath Volume		0.882	01:23	
Air Blank		0.000	01:24	
Control Test		0.077	01:24	
Air Blank		0.000	01:25	
Diagnostics Check		OK	01:25	

Data Download Date/Time: 02/08/2017 19:31

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRIMBLE,TERESA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:44

SUBJECT'S LAST NAME: GREIG

DATE/TIME OF TEST: 01/27/2017 01:35:33

Diagnostics Check		OK	01:27
Air Blank		0.000	01:28
Control Test		0.077	01:28
Air Blank		0.000	01:29
Subject Sample	# 1	0.149	01:29
Breath Volume		1.375	01:29
Air Blank		0.000	01:30
Air Blank		0.000	01:32
Subject Sample	# 2	0.143	01:33
Breath Volume		1.546	01:33
Air Blank		0.000	01:34
Control Test		0.078	01:34
Air Blank		0.000	01:35
Diagnostics Check		OK	01:35

Data Download Date/Time: 02/08/2017 19:31

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY,PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MANGANARO,J
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:30

SUBJECT'S LAST NAME: SEAVEY

DATE/TIME OF TEST: 01/27/2017 05:57:22

Diagnostics Check		OK	05:50
Air Blank		0.000	05:51
Control Test		0.079	05:51
Air Blank		0.000	05:52
Subject Sample	# 1	0.181	05:52
Breath Volume		1.609	05:52
Air Blank		0.000	05:53
Air Blank		0.000	05:55
Subject Sample	# 2	0.171	05:55
Breath Volume		1.390	05:55
Air Blank		0.000	05:56
Control Test		0.078	05:56
Air Blank		0.000	05:57
Diagnostics Check		OK	05:57

Data Download Date/Time: 02/08/2017 19:31

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: COSTA,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: TORRES

DATE/TIME OF TEST: 01/28/2017 02:16:21

Diagnostics Check		OK	02:10
Air Blank		0.000	02:10
Control Test		0.080	02:11
Air Blank		0.000	02:11
Subject Sample	# 1	0.000	02:12
Breath Volume		2.281	02:12
Air Blank		0.000	02:12
Air Blank		0.000	02:14
Subject Sample	# 2	0.000	02:14
Breath Volume		2.050	02:14
Air Blank		0.000	02:15
Control Test		0.080	02:15
Air Blank		0.000	02:16
Diagnostics Check		OK	02:16

Data Download Date/Time: 02/08/2017 19:31

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY,PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PECYLAK,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:58

SUBJECT'S LAST NAME: LOVE

DATE/TIME OF TEST: 01/28/2017 04:26:54

Diagnostics Check		OK	04:20
Air Blank		0.000	04:20
Control Test		0.079	04:21
Air Blank		0.000	04:21
Subject Sample	# 1	0.000	04:22
Breath Volume		2.125	04:22
Air Blank		0.000	04:23
Air Blank		0.000	04:25
Subject Sample	# 2	0.000	04:25
Breath Volume		1.937	04:25
Air Blank		0.000	04:25
Control Test		0.079	04:26
Air Blank		0.000	04:26
Diagnostics Check		OK	04:26

Data Download Date/Time: 02/08/2017 19:31

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,REBECCA,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DEVOE,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 15:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:24

SUBJECT'S LAST NAME: MOTILALL

DATE/TIME OF TEST: 01/28/2017 17:01:08

Diagnostics Check		OK	16:49	
Air Blank		0.000	16:49	
Control Test		0.080	16:50	
Air Blank		0.000	16:50	
Subject Sample	# 1	0.198	16:53	Volume Not Met
Breath Volume		0.179	16:53	
Air Blank		0.000	16:54	
Air Blank		0.000	16:56	
Subject Sample	# 2	0.191	16:59	Volume Not Met
Breath Volume		0.054	16:59	
Air Blank		0.000	17:00	
Control Test		0.077	17:00	
Air Blank		0.000	17:01	
Diagnostics Check		OK	17:01	

Data Download Date/Time: 02/08/2017 19:31

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,REBECCA,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DEVOE,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 15:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:24

SUBJECT'S LAST NAME: MOTILALL

DATE/TIME OF TEST: 01/28/2017 17:15:54

Diagnostics Check		OK	17:07
Air Blank		0.000	17:08
Control Test		0.077	17:08
Air Blank		0.000	17:09
Subject Sample	# 1	0.224	17:11
Breath Volume		1.339	17:11
Air Blank		0.000	17:12
Air Blank		0.000	17:13
Subject Sample	# 2	0.223	17:14
Breath Volume		1.593	17:14
Air Blank		0.000	17:14
Control Test		0.077	17:15
Air Blank		0.000	17:15
Diagnostics Check		OK	17:15

Data Download Date/Time: 02/08/2017 19:31

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PORTMAN,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 19:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:40

SUBJECT'S LAST NAME: WELTY

DATE/TIME OF TEST: 01/28/2017 20:08:23

Diagnostics Check		OK	20:01
Air Blank		0.000	20:01
Control Test		0.079	20:01
Air Blank		0.000	20:02
Subject Sample	# 1	0.111	20:03
Breath Volume		1.445	20:03
Air Blank		0.000	20:04
Air Blank		0.000	20:05
Subject Sample	# 2	0.105	20:06
Breath Volume		1.734	20:06
Air Blank		0.000	20:07
Control Test		0.079	20:07
Air Blank		0.000	20:08
Diagnostics Check		OK	20:08

Data Download Date/Time: 02/08/2017 19:31

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:58

SUBJECT'S LAST NAME: TAMAYO

DATE/TIME OF TEST: 01/29/2017 00:24:58

Diagnostics Check		OK	00:19
Air Blank		0.000	00:19
Control Test		0.080	00:19
Air Blank		0.000	00:20
Subject Sample	# 1	0.000	00:20
Breath Volume		1.480	00:20
Air Blank		0.000	00:21
Air Blank		0.000	00:23
Subject Sample	# 2	0.000	00:23
Breath Volume		2.019	00:23
Air Blank		0.000	00:24
Control Test		0.080	00:24
Air Blank		0.000	00:24
Diagnostics Check		OK	00:24

Data Download Date/Time: 02/08/2017 19:31

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 05:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:31

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 01/29/2017 05:58:28

Diagnostics Check		OK	05:51
Air Blank		0.000	05:52
Control Test		0.079	05:52
Air Blank		0.000	05:53
Subject Sample	# 1	0.158	05:53
Breath Volume		1.726	05:53
Air Blank		0.000	05:54
Air Blank		0.000	05:56
Subject Sample	# 2	0.158	05:56
Breath Volume		1.535	05:56
Air Blank		0.000	05:57
Control Test		0.078	05:57
Air Blank		0.000	05:58
Diagnostics Check		OK	05:58

Data Download Date/Time: 02/08/2017 19:31

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: AGUAYO,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 17:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:36

SUBJECT'S LAST NAME: BRUECK

DATE/TIME OF TEST: 01/29/2017 19:03:54

Diagnostics Check		OK	18:56
Air Blank		0.000	18:57
Control Test		0.080	18:57
Air Blank		0.000	18:57
Subject Sample	# 1	0.187	18:59
Breath Volume		1.988	18:59
Air Blank		0.000	19:00
Air Blank		0.000	19:01
Subject Sample	# 2	0.192	19:02
Breath Volume		1.519	19:02
Air Blank		0.000	19:02
Control Test		0.077	19:03
Air Blank		0.000	19:03
Diagnostics Check		OK	19:03

Data Download Date/Time: 02/08/2017 19:31

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY,PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MCLELLAND,E
ARRESTING OFFICER AGENCY: Macdill Air Force Ba
LAST AGENCY INSPECTION: 01/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 19:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:07

SUBJECT'S LAST NAME: CORDARO

DATE/TIME OF TEST: 01/31/2017 20:36:46

Diagnostics Check		OK	20:29
Air Blank		0.000	20:30
Control Test		0.080	20:30
Air Blank		0.000	20:31
Subject Sample	# 1	0.000	20:32
Breath Volume		3.046	20:32
Air Blank		0.000	20:32
Air Blank		0.000	20:34
Subject Sample	# 2	0.000	20:35
Breath Volume		2.960	20:35
Air Blank		0.000	20:35
Control Test		0.080	20:36
Air Blank		0.000	20:36
Diagnostics Check		OK	20:36

Data Download Date/Time: 02/08/2017 19:31

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRIMBLE,TERESA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRONSON,R
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:42

SUBJECT'S LAST NAME: TROCHE

DATE/TIME OF TEST: 01/31/2017 23:08:54

Diagnostics Check		OK	23:02
Air Blank		0.000	23:03
Control Test		0.079	23:03
Air Blank		0.000	23:03
Subject Sample	# 1	0.180	23:04
Breath Volume		1.441	23:04
Air Blank		0.000	23:05
Air Blank		0.000	23:06
Subject Sample	# 2	0.189	23:07
Breath Volume		1.472	23:07
Air Blank		0.000	23:07
Control Test		0.078	23:08
Air Blank		0.000	23:08
Diagnostics Check		OK	23:08

Data Download Date/Time: 02/08/2017 19:31

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: AGUAYO,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:27

SUBJECT'S LAST NAME: GAGNER

DATE/TIME OF TEST: 01/01/2017 00:54:08

Diagnostics Check		OK	00:47
Air Blank		0.000	00:48
Control Test		0.081	00:48
Air Blank		0.000	00:49
Subject Sample	# 1	0.162	00:49
Breath Volume		1.566	00:49
Air Blank		0.000	00:50
Air Blank		0.000	00:52
Subject Sample	# 2	0.162	00:52
Breath Volume		1.574	00:52
Air Blank		0.000	00:53
Control Test		0.080	00:53
Air Blank		0.000	00:54
Diagnostics Check		OK	00:54

Data Download Date/Time: 01/10/2017 19:52

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LARA,A
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:29

SUBJECT'S LAST NAME: ALKABI

DATE/TIME OF TEST: 01/01/2017 04:59:24

Diagnostics Check		OK	04:49
Air Blank		0.000	04:49
Control Test		0.081	04:50
Air Blank		0.000	04:50
Subject Sample	# 1	0.144	04:51
Breath Volume		1.601	04:51
Air Blank		0.000	04:52
Air Blank		0.000	04:54
Subject Sample	# 2	0.170	04:54
Breath Volume		2.402	04:54
Air Blank		0.000	04:55
Air Blank		0.000	04:57
Subject Sample	# 3	0.162	04:57
Breath Volume		1.695	04:57
Air Blank		0.000	04:58
Control Test		0.081	04:58
Air Blank		0.000	04:59
Diagnostics Check		OK	04:59

Data Download Date/Time: 01/10/2017 19:52

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK,CASEY,S
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: NIEUWENDAM,R
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 12/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:19

SUBJECT'S LAST NAME: CESPEDES DIAZ

DATE/TIME OF TEST: 01/02/2017 00:50:00

Diagnostics Check		OK	00:41
Air Blank		0.000	00:41
Control Test		0.081	00:42
Air Blank		0.000	00:42
Subject Sample	# 1	0.175	00:43
Breath Volume		1.843	00:43
Air Blank		0.000	00:44
Air Blank		0.000	00:46
Subject Sample	# 2	0.176	00:48
Breath Volume		2.078	00:48
Air Blank		0.000	00:49
Control Test		0.080	00:49
Air Blank		0.000	00:49
Diagnostics Check		OK	00:50

Data Download Date/Time: 01/10/2017 19:52

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK,CASEY,S
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: NIEUWENDAM,R
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 12/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:52

SUBJECT'S LAST NAME: CESPEDES DIAZ

DATE/TIME OF TEST: 01/02/2017 01:26:56

Diagnostics Check		OK	01:13	
Air Blank		0.000	01:14	
Control Test		0.080	01:14	
Air Blank		0.000	01:14	
Subject Sample	# 1	0.160	01:15	
Breath Volume		1.210	01:15	
Air Blank		0.000	01:16	
Air Blank		0.000	01:17	
Subject Sample	# 2		01:21	No Sample Provided
Breath Volume		0.000	01:21	
Air Blank		0.000	01:21	
Air Blank		0.000	01:23	
Subject Sample	# 3	0.152	01:25	
Breath Volume		1.148	01:25	
Air Blank		0.000	01:26	
Control Test		0.080	01:26	
Air Blank		0.000	01:26	
Diagnostics Check		OK	01:26	

Data Download Date/Time: 01/10/2017 19:52

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 12/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:01

SUBJECT'S LAST NAME: GUSTIN

DATE/TIME OF TEST: 01/02/2017 02:28:28

Diagnostics Check		OK	02:21
Air Blank		0.000	02:22
Control Test		0.081	02:22
Air Blank		0.000	02:23
Subject Sample	# 1	0.141	02:24
Breath Volume		3.167	02:24
Air Blank		0.000	02:24
Air Blank		0.000	02:26
Subject Sample	# 2	0.134	02:26
Breath Volume		2.945	02:26
Air Blank		0.000	02:27
Control Test		0.080	02:27
Air Blank		0.000	02:28
Diagnostics Check		OK	02:28

Data Download Date/Time: 01/10/2017 19:52

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 12/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:54

SUBJECT'S LAST NAME: RORECH

DATE/TIME OF TEST: 01/03/2017 00:20:35

Diagnostics Check		OK	00:14
Air Blank		0.000	00:14
Control Test		0.081	00:15
Air Blank		0.000	00:15
Subject Sample	# 1	0.000	00:16
Breath Volume		3.875	00:16
Air Blank		0.000	00:16
Air Blank		0.000	00:18
Subject Sample	# 2	0.000	00:19
Breath Volume		3.128	00:19
Air Blank		0.000	00:19
Control Test		0.081	00:19
Air Blank		0.000	00:20
Diagnostics Check		OK	00:20

Data Download Date/Time: 01/10/2017 19:52

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK,CASEY,S
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: NIEUWENDAM,R
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 12/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:24

SUBJECT'S LAST NAME: FISHER

DATE/TIME OF TEST: 01/03/2017 05:49:44

Diagnostics Check		OK	05:44	
Air Blank		0.000	05:45	
Control Test		0.081	05:45	
Air Blank		0.000	05:45	
Subject Sample	# 1		05:48	Subject Test Refused
Breath Volume		0.000	05:48	
Air Blank		0.000	05:48	
Control Test		0.081	05:49	
Air Blank		0.000	05:49	
Diagnostics Check		OK	05:49	

Data Download Date/Time: 01/10/2017 19:52

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 12/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:04

SUBJECT'S LAST NAME: MANIGO

DATE/TIME OF TEST: 01/03/2017 23:33:13

Diagnostics Check		OK	23:24	
Air Blank		0.000	23:25	
Control Test		0.081	23:25	
Air Blank		0.000	23:26	
Subject Sample	# 1		23:29	No Sample Provided
Breath Volume		0.000	23:29	
Air Blank		0.000	23:29	
Air Blank		0.000	23:31	
Subject Sample	# 2		23:31	Subject Test Refused
Breath Volume		0.000	23:31	
Air Blank		0.000	23:32	
Control Test		0.081	23:32	
Air Blank		0.000	23:33	
Diagnostics Check		OK	23:33	

Data Download Date/Time: 01/10/2017 19:52

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRIMBLE,TERESA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRONSON,R
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:09

SUBJECT'S LAST NAME: BIERLEY

DATE/TIME OF TEST: 01/04/2017 00:36:22

Diagnostics Check		OK	00:29
Air Blank		0.000	00:30
Control Test		0.080	00:30
Air Blank		0.000	00:30
Subject Sample	# 1	0.164	00:31
Breath Volume		3.894	00:31
Air Blank		0.000	00:32
Air Blank		0.000	00:34
Subject Sample	# 2	0.166	00:34
Breath Volume		3.664	00:34
Air Blank		0.000	00:35
Control Test		0.080	00:35
Air Blank		0.000	00:36
Diagnostics Check		OK	00:36

Data Download Date/Time: 01/10/2017 19:52

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKIPPER,ROGER,G
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BURKE,B 253503
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 12:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:17

SUBJECT'S LAST NAME: SESSLER

DATE/TIME OF TEST: 01/04/2017 13:46:09

Diagnostics Check		OK	13:39
Air Blank		0.000	13:40
Control Test		0.080	13:40
Air Blank		0.000	13:40
Subject Sample	# 1	0.109	13:41
Breath Volume		1.824	13:41
Air Blank		0.000	13:42
Air Blank		0.000	13:44
Subject Sample	# 2	0.109	13:44
Breath Volume		1.445	13:44
Air Blank		0.000	13:45
Control Test		0.080	13:45
Air Blank		0.000	13:46
Diagnostics Check		OK	13:46

Data Download Date/Time: 01/10/2017 19:52

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GONZALEZ,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 12/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:42

SUBJECT'S LAST NAME: MASOUD

DATE/TIME OF TEST: 01/05/2017 04:09:27

Diagnostics Check		OK	04:03
Air Blank		0.000	04:03
Control Test		0.081	04:04
Air Blank		0.000	04:04
Subject Sample	# 1	0.068	04:05
Breath Volume		2.644	04:05
Air Blank		0.000	04:05
Air Blank		0.000	04:07
Subject Sample	# 2	0.067	04:07
Breath Volume		1.316	04:07
Air Blank		0.000	04:08
Control Test		0.081	04:08
Air Blank		0.000	04:09
Diagnostics Check		OK	04:09

Data Download Date/Time: 01/10/2017 19:52

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKIPPER,ROGER,G
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HAZELZET,D 550
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 12/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 07:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:00

SUBJECT'S LAST NAME: YOUNGBLOOD

DATE/TIME OF TEST: 01/05/2017 08:29:56

Diagnostics Check		OK	08:22
Air Blank		0.000	08:22
Control Test		0.082	08:23
Air Blank		0.000	08:23
Subject Sample	# 1	0.000	08:24
Breath Volume		1.734	08:24
Air Blank		0.000	08:25
Air Blank		0.000	08:27
Subject Sample	# 2	0.000	08:28
Breath Volume		1.839	08:28
Air Blank		0.000	08:29
Control Test		0.080	08:29
Air Blank		0.000	08:29
Diagnostics Check		OK	08:29

Data Download Date/Time: 01/10/2017 19:52

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRENNAN,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: ZEFERINO DE JESUS

DATE/TIME OF TEST: 01/07/2017 02:02:08

Diagnostics Check		OK	01:55
Air Blank		0.000	01:55
Control Test		0.081	01:56
Air Blank		0.000	01:56
Subject Sample	# 1	0.203	01:57
Breath Volume		1.796	01:57
Air Blank		0.000	01:58
Air Blank		0.000	01:59
Subject Sample	# 2	0.217	02:00
Breath Volume		2.648	02:00
Air Blank		0.000	02:01
Control Test		0.080	02:01
Air Blank		0.000	02:02
Diagnostics Check		OK	02:02

Data Download Date/Time: 01/10/2017 19:52

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRABAND,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 12/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:03

SUBJECT'S LAST NAME: NEWELL

DATE/TIME OF TEST: 01/07/2017 03:30:44

Diagnostics Check		OK	03:23
Air Blank		0.000	03:24
Control Test		0.080	03:24
Air Blank		0.000	03:25
Subject Sample	# 1	0.155	03:26
Breath Volume		2.308	03:26
Air Blank		0.000	03:26
Air Blank		0.000	03:28
Subject Sample	# 2	0.161	03:29
Breath Volume		2.246	03:29
Air Blank		0.000	03:29
Control Test		0.080	03:30
Air Blank		0.000	03:30
Diagnostics Check		OK	03:30

Data Download Date/Time: 01/10/2017 19:52

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK,CASEY,S
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DEJESUS,V
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:47

SUBJECT'S LAST NAME: NOLASCO CARBAJAL

DATE/TIME OF TEST: 01/07/2017 04:13:59

Diagnostics Check		OK	04:07
Air Blank		0.000	04:07
Control Test		0.082	04:08
Air Blank		0.000	04:08
Subject Sample	# 1	0.241	04:09
Breath Volume		2.238	04:09
Air Blank		0.000	04:10
Air Blank		0.000	04:11
Subject Sample	# 2	0.248	04:12
Breath Volume		2.492	04:12
Air Blank		0.000	04:13
Control Test		0.080	04:13
Air Blank		0.000	04:13
Diagnostics Check		OK	04:13

Data Download Date/Time: 01/10/2017 19:52

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FRANKLIN,R
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 05:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:12

SUBJECT'S LAST NAME: CHANDLER

DATE/TIME OF TEST: 01/07/2017 06:39:13

Diagnostics Check		OK	06:32
Air Blank		0.000	06:33
Control Test		0.080	06:33
Air Blank		0.000	06:33
Subject Sample	# 1	0.174	06:34
Breath Volume		1.226	06:34
Air Blank		0.000	06:34
Air Blank		0.000	06:36
Subject Sample	# 2	0.171	06:37
Breath Volume		1.710	06:37
Air Blank		0.000	06:38
Control Test		0.080	06:38
Air Blank		0.000	06:39
Diagnostics Check		OK	06:39

Data Download Date/Time: 01/10/2017 19:52

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GONZALEZ,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:41

SUBJECT'S LAST NAME: DALFINO

DATE/TIME OF TEST: 01/07/2017 23:08:47

Diagnostics Check		OK	23:02
Air Blank		0.000	23:02
Control Test		0.082	23:02
Air Blank		0.000	23:03
Subject Sample	# 1	0.233	23:04
Breath Volume		2.648	23:04
Air Blank		0.000	23:04
Air Blank		0.000	23:06
Subject Sample	# 2	0.218	23:07
Breath Volume		1.433	23:07
Air Blank		0.000	23:07
Control Test		0.080	23:08
Air Blank		0.000	23:08
Diagnostics Check		OK	23:08

Data Download Date/Time: 01/10/2017 19:52

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY,PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LYON,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 12/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: LUCAS

DATE/TIME OF TEST: 01/08/2017 01:42:28

Diagnostics Check		OK	01:35
Air Blank		0.000	01:36
Control Test		0.081	01:36
Air Blank		0.000	01:37
Subject Sample	# 1	0.182	01:38
Breath Volume		1.457	01:38
Air Blank		0.000	01:38
Air Blank		0.000	01:40
Subject Sample	# 2	0.184	01:40
Breath Volume		2.050	01:40
Air Blank		0.000	01:41
Control Test		0.081	01:41
Air Blank		0.000	01:42
Diagnostics Check		OK	01:42

Data Download Date/Time: 01/10/2017 19:52

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRIMBLE,TERESA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DEJESUS,V
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:21

SUBJECT'S LAST NAME: LEWIS

DATE/TIME OF TEST: 01/08/2017 05:49:05

Diagnostics Check		OK	05:42
Air Blank		0.000	05:43
Control Test		0.082	05:43
Air Blank		0.000	05:44
Subject Sample	# 1	0.132	05:44
Breath Volume		2.273	05:44
Air Blank		0.000	05:45
Air Blank		0.000	05:46
Subject Sample	# 2	0.129	05:47
Breath Volume		3.300	05:47
Air Blank		0.000	05:48
Control Test		0.081	05:48
Air Blank		0.000	05:49
Diagnostics Check		OK	05:49

Data Download Date/Time: 01/10/2017 19:52

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 12/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:44

SUBJECT'S LAST NAME: AMATO

DATE/TIME OF TEST: 01/08/2017 23:11:03

Diagnostics Check		OK	23:04
Air Blank		0.000	23:05
Control Test		0.082	23:05
Air Blank		0.000	23:06
Subject Sample	# 1	0.138	23:06
Breath Volume		3.355	23:06
Air Blank		0.000	23:07
Air Blank		0.000	23:09
Subject Sample	# 2	0.147	23:09
Breath Volume		3.269	23:09
Air Blank		0.000	23:10
Control Test		0.082	23:10
Air Blank		0.000	23:10
Diagnostics Check		OK	23:11

Data Download Date/Time: 01/10/2017 19:52

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARLAUG,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 12/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:57

SUBJECT'S LAST NAME: VELEZ

DATE/TIME OF TEST: 01/09/2017 00:30:05

Diagnostics Check		OK	00:17	
Air Blank		0.000	00:18	
Control Test		0.082	00:18	
Air Blank		0.000	00:18	
Subject Sample	# 1	0.167	00:22	Volume Not Met
Breath Volume		0.832	00:22	
Air Blank		0.000	00:22	
Air Blank		0.000	00:24	
Subject Sample	# 2	0.197	00:25	
Breath Volume		1.222	00:25	
Air Blank		0.000	00:26	
Air Blank		0.000	00:27	
Subject Sample	# 3	0.192	00:28	
Breath Volume		2.085	00:28	
Air Blank		0.000	00:29	
Control Test		0.081	00:29	
Air Blank		0.000	00:30	
Diagnostics Check		OK	00:30	

Data Download Date/Time: 01/10/2017 19:52

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRIMBLE,TERESA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 12/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:41

SUBJECT'S LAST NAME: PAIGE-PARKS

DATE/TIME OF TEST: 01/10/2017 04:08:23

Diagnostics Check		OK	04:01
Air Blank		0.000	04:02
Control Test		0.082	04:02
Air Blank		0.000	04:03
Subject Sample	# 1	0.125	04:03
Breath Volume		3.746	04:03
Air Blank		0.000	04:04
Air Blank		0.000	04:06
Subject Sample	# 2	0.120	04:06
Breath Volume		3.171	04:06
Air Blank		0.000	04:07
Control Test		0.081	04:07
Air Blank		0.000	04:08
Diagnostics Check		OK	04:08

Data Download Date/Time: 01/10/2017 19:52

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: AGUAYO,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:04

SUBJECT'S LAST NAME: GARCIA FIGUEROA

DATE/TIME OF TEST: 01/28/2017 02:31:52

Diagnostics Check		OK	02:24
Air Blank		0.000	02:25
Control Test		0.082	02:25
Air Blank		0.000	02:25
Subject Sample	# 1	0.082	02:27
Breath Volume		1.945	02:27
Air Blank		0.000	02:27
Air Blank		0.000	02:29
Subject Sample	# 2	0.085	02:30
Breath Volume		1.136	02:30
Air Blank		0.000	02:30
Control Test		0.081	02:31
Air Blank		0.000	02:31
Diagnostics Check		OK	02:31

Data Download Date/Time: 02/08/2017 19:16

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: NIEUWENDAM,R
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:11

SUBJECT'S LAST NAME: SHACKELFORD

DATE/TIME OF TEST: 01/28/2017 21:40:33

Diagnostics Check		OK	21:34
Air Blank		0.000	21:34
Control Test		0.082	21:35
Air Blank		0.000	21:35
Subject Sample	# 1	0.136	21:36
Breath Volume		2.460	21:36
Air Blank		0.000	21:36
Air Blank		0.000	21:38
Subject Sample	# 2	0.138	21:39
Breath Volume		2.007	21:39
Air Blank		0.000	21:39
Control Test		0.081	21:39
Air Blank		0.000	21:40
Diagnostics Check		OK	21:40

Data Download Date/Time: 02/08/2017 19:16

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK,CASEY,S
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARLAUG,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:14

SUBJECT'S LAST NAME: ALMBERG

DATE/TIME OF TEST: 01/29/2017 00:42:18

Diagnostics Check		OK	00:35
Air Blank		0.000	00:36
Control Test		0.082	00:36
Air Blank		0.000	00:37
Subject Sample	# 1	0.232	00:37
Breath Volume		3.554	00:37
Air Blank		0.000	00:38
Air Blank		0.000	00:40
Subject Sample	# 2	0.233	00:40
Breath Volume		3.343	00:40
Air Blank		0.000	00:41
Control Test		0.081	00:41
Air Blank		0.000	00:42
Diagnostics Check		OK	00:42

Data Download Date/Time: 02/08/2017 19:16

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK,CASEY,S
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: NIEUWENDAM,R
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:36

SUBJECT'S LAST NAME: BARNES JR

DATE/TIME OF TEST: 01/29/2017 02:09:00

Diagnostics Check		OK	01:58
Air Blank		0.000	01:59
Control Test		0.081	01:59
Air Blank		0.000	02:00
Subject Sample	# 1	0.145	02:01
Breath Volume		3.742	02:01
Air Blank		0.000	02:02
Air Blank		0.000	02:04
Subject Sample	# 2	0.123	02:04
Breath Volume		1.832	02:04
Air Blank		0.000	02:05
Air Blank		0.000	02:07
Subject Sample	# 3	0.131	02:07
Breath Volume		1.207	02:07
Air Blank		0.000	02:08
Control Test		0.081	02:08
Air Blank		0.000	02:08
Diagnostics Check		OK	02:09

Data Download Date/Time: 02/08/2017 19:16

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GONZALEZ,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:03

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 01/29/2017 02:30:43

Diagnostics Check		OK	02:23
Air Blank		0.000	02:24
Control Test		0.081	02:24
Air Blank		0.000	02:24
Subject Sample	# 1	0.116	02:25
Breath Volume		1.429	02:25
Air Blank		0.000	02:26
Air Blank		0.000	02:28
Subject Sample	# 2	0.117	02:29
Breath Volume		1.589	02:29
Air Blank		0.000	02:29
Control Test		0.081	02:30
Air Blank		0.000	02:30
Diagnostics Check		OK	02:30

Data Download Date/Time: 02/08/2017 19:16

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK,CASEY,S
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VAN-TREESE,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:48

SUBJECT'S LAST NAME: STOCKER

DATE/TIME OF TEST: 01/29/2017 04:18:40

Diagnostics Check		OK	04:09
Air Blank		0.000	04:09
Control Test		0.081	04:09
Air Blank		0.000	04:10
Subject Sample	# 1	0.147	04:12
Breath Volume		1.117	04:12
Air Blank		0.000	04:13
Air Blank		0.000	04:15
Subject Sample	# 2	0.149	04:17
Breath Volume		1.144	04:17
Air Blank		0.000	04:17
Control Test		0.080	04:18
Air Blank		0.000	04:18
Diagnostics Check		OK	04:18

Data Download Date/Time: 02/08/2017 19:16

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DUNCAN,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:19

SUBJECT'S LAST NAME: MERZ

DATE/TIME OF TEST: 01/29/2017 21:46:33

Diagnostics Check		OK	21:39
Air Blank		0.000	21:40
Control Test		0.081	21:40
Air Blank		0.000	21:40
Subject Sample	# 1	0.289	21:41
Breath Volume		3.058	21:41
Air Blank		0.000	21:42
Air Blank		0.000	21:44
Subject Sample	# 2	0.292	21:44
Breath Volume		5.402	21:44
Air Blank		0.000	21:45
Control Test		0.080	21:45
Air Blank		0.000	21:46
Diagnostics Check		OK	21:46

Data Download Date/Time: 02/08/2017 19:16

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:34

SUBJECT'S LAST NAME: JORDAN-BENITEZ

DATE/TIME OF TEST: 01/30/2017 00:00:43

Diagnostics Check		OK	23:54
Air Blank		0.000	23:54
Control Test		0.082	23:55
Air Blank		0.000	23:55
Subject Sample	# 1	0.144	23:56
Breath Volume		1.519	23:56
Air Blank		0.000	23:56
Air Blank		0.000	23:58
Subject Sample	# 2	0.144	23:59
Breath Volume		1.417	23:59
Air Blank		0.000	23:59
Control Test		0.080	00:00
Air Blank		0.000	00:00
Diagnostics Check		OK	00:00

Data Download Date/Time: 02/08/2017 19:16

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK,CASEY,S
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRENNAN,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:32

SUBJECT'S LAST NAME: NULL

DATE/TIME OF TEST: 01/30/2017 23:00:47

Diagnostics Check		OK	22:54
Air Blank		0.000	22:55
Control Test		0.083	22:55
Air Blank		0.000	22:55
Subject Sample	# 1	0.178	22:56
Breath Volume		2.785	22:56
Air Blank		0.000	22:57
Air Blank		0.000	22:58
Subject Sample	# 2	0.177	22:59
Breath Volume		2.542	22:59
Air Blank		0.000	22:59
Control Test		0.081	23:00
Air Blank		0.000	23:00
Diagnostics Check		OK	23:00

Data Download Date/Time: 02/08/2017 19:16

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK,CASEY,S
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARLAUG,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:33

SUBJECT'S LAST NAME: POTTS III

DATE/TIME OF TEST: 01/31/2017 03:01:08

Diagnostics Check		OK	02:54
Air Blank		0.000	02:54
Control Test		0.083	02:54
Air Blank		0.000	02:55
Subject Sample	# 1	0.125	02:56
Breath Volume		2.031	02:56
Air Blank		0.000	02:56
Air Blank		0.000	02:58
Subject Sample	# 2	0.125	02:59
Breath Volume		1.945	02:59
Air Blank		0.000	03:00
Control Test		0.083	03:00
Air Blank		0.000	03:01
Diagnostics Check		OK	03:01

Data Download Date/Time: 02/08/2017 19:16

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRABAND,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:40

SUBJECT'S LAST NAME: DYER

DATE/TIME OF TEST: 01/31/2017 05:06:42

Diagnostics Check		OK	05:00
Air Blank		0.000	05:00
Control Test		0.082	05:01
Air Blank		0.000	05:01
Subject Sample	# 1	0.155	05:02
Breath Volume		1.820	05:02
Air Blank		0.000	05:02
Air Blank		0.000	05:04
Subject Sample	# 2	0.156	05:05
Breath Volume		2.042	05:05
Air Blank		0.000	05:05
Control Test		0.082	05:06
Air Blank		0.000	05:06
Diagnostics Check		OK	05:06

Data Download Date/Time: 02/08/2017 19:16

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARLAUG,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: DANIELS

DATE/TIME OF TEST: 01/31/2017 23:57:02

Diagnostics Check		OK	23:50
Air Blank		0.000	23:51
Control Test		0.082	23:51
Air Blank		0.000	23:52
Subject Sample	# 1	0.106	23:52
Breath Volume		3.000	23:52
Air Blank		0.000	23:53
Air Blank		0.000	23:54
Subject Sample	# 2	0.104	23:55
Breath Volume		3.039	23:55
Air Blank		0.000	23:56
Control Test		0.081	23:56
Air Blank		0.000	23:56
Diagnostics Check		OK	23:57

Data Download Date/Time: 02/08/2017 19:16

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003406

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HATCHER,DUSTIN,D
BREATH TEST OPERATOR AGENCY: Leon County SO
ARRESTING OFFICER: JEFFERSON-SHAW,PATRICIA
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 12/06/2016

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 01/05/2017 01:18:18

Diagnostics Check		OK	01:11
Air Blank		0.000	01:11
Control Test		0.081	01:11
Air Blank		0.000	01:12
Subject Sample	# 1	0.131	01:13
Breath Volume		1.386	01:13
Air Blank		0.000	01:14
Air Blank		0.000	01:16
Subject Sample	# 2	0.137	01:16
Breath Volume		1.453	01:16
Air Blank		0.000	01:17
Control Test		0.080	01:17
Air Blank		0.000	01:18
Diagnostics Check		OK	01:18

Data Download Date/Time: 01/09/2017 11:10

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003411

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANGELICCHI,JOSEPH,S
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: ANGELICCHI,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 477627
DRY GAS EXPIRY DATE: 06/29/2019
VIOLATION CODE: DUI
ARREST TIME: 00:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: RUSSELL

DATE/TIME OF TEST: 01/01/2017 01:10:35

Diagnostics Check		OK	01:04
Air Blank		0.000	01:04
Control Test		0.080	01:05
Air Blank		0.000	01:05
Subject Sample	# 1	0.109	01:06
Breath Volume		2.273	01:06
Air Blank		0.000	01:06
Air Blank		0.000	01:08
Subject Sample	# 2	0.105	01:09
Breath Volume		2.628	01:09
Air Blank		0.000	01:09
Control Test		0.079	01:09
Air Blank		0.000	01:10
Diagnostics Check		OK	01:10

Data Download Date/Time: 01/26/2017 00:47

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003411

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANGELICCHI,JOSEPH,S
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: ANGELICCHI,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 477627
DRY GAS EXPIRY DATE: 06/29/2019
VIOLATION CODE: DUI
ARREST TIME: 03:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: ABUDAYEH

DATE/TIME OF TEST: 01/07/2017 04:02:22

Diagnostics Check		OK	03:55
Air Blank		0.000	03:56
Control Test		0.079	03:56
Air Blank		0.000	03:57
Subject Sample	# 1	0.101	03:57
Breath Volume		3.648	03:57
Air Blank		0.000	03:58
Air Blank		0.000	04:00
Subject Sample	# 2	0.103	04:00
Breath Volume		3.359	04:00
Air Blank		0.000	04:01
Control Test		0.079	04:01
Air Blank		0.000	04:02
Diagnostics Check		OK	04:02

Data Download Date/Time: 01/26/2017 00:47

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003411

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANGELICCHI,JOSEPH,S
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: ANGELICCHI,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 477627
DRY GAS EXPIRY DATE: 06/29/2019
VIOLATION CODE: DUI
ARREST TIME: 20:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:35

SUBJECT'S LAST NAME: BAEZ

DATE/TIME OF TEST: 01/07/2017 21:04:25

Diagnostics Check		OK	20:58
Air Blank		0.000	20:58
Control Test		0.080	20:59
Air Blank		0.000	20:59
Subject Sample	# 1	0.019	21:00
Breath Volume		2.304	21:00
Air Blank		0.000	21:00
Air Blank		0.000	21:02
Subject Sample	# 2	0.019	21:02
Breath Volume		3.402	21:02
Air Blank		0.000	21:03
Control Test		0.081	21:03
Air Blank		0.000	21:04
Diagnostics Check		OK	21:04

Data Download Date/Time: 01/26/2017 00:47

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003411

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPBELL,JOSHUA,C
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: CAMPBELL,JOSHUA
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 477627
DRY GAS EXPIRY DATE: 06/29/2019
VIOLATION CODE: DUI
ARREST TIME: 00:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: BRACE

DATE/TIME OF TEST: 01/08/2017 00:57:58

Diagnostics Check		OK	00:55	
Air Blank		0.000	00:56	
Control Test		0.080	00:56	
Air Blank		0.000	00:57	
Subject Sample	# 1		00:57	RFI Detect
Breath Volume		0.000	00:57	
Air Blank			00:57	RFI Detect
Air Blank			00:57	RFI Detect

Data Download Date/Time: 01/26/2017 00:47

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003411

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPBELL,JOSHUA,C
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: CAMPBELL,JOSHUA
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 477627
DRY GAS EXPIRY DATE: 06/29/2019
VIOLATION CODE: DUI
ARREST TIME: 00:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: BRACE

DATE/TIME OF TEST: 01/08/2017 01:07:50

Diagnostics Check		OK	00:59
Air Blank		0.000	00:59
Control Test		0.080	01:00
Air Blank		0.000	01:00
Subject Sample	# 1	0.077	01:01
Breath Volume		2.507	01:01
Air Blank		0.000	01:01
Air Blank		0.000	01:03
Subject Sample	# 2	0.077	01:06
Breath Volume		1.320	01:06
Air Blank		0.000	01:06
Control Test		0.081	01:07
Air Blank		0.000	01:07
Diagnostics Check		OK	01:07

Data Download Date/Time: 01/26/2017 00:47

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003411

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPBELL,JOSHUA,C
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: CAMPBELL,JOSHUA
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 477627
DRY GAS EXPIRY DATE: 06/29/2019
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: DAVILA

DATE/TIME OF TEST: 01/09/2017 03:08:45

Diagnostics Check		OK	03:02
Air Blank		0.000	03:02
Control Test		0.081	03:02
Air Blank		0.000	03:03
Subject Sample	# 1	0.092	03:04
Breath Volume		2.359	03:04
Air Blank		0.000	03:04
Air Blank		0.000	03:06
Subject Sample	# 2	0.090	03:07
Breath Volume		1.824	03:07
Air Blank		0.000	03:07
Control Test		0.081	03:08
Air Blank		0.000	03:08
Diagnostics Check		OK	03:08

Data Download Date/Time: 01/26/2017 00:47

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003411

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,JASON,G
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: BROWN,JASON
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 477627
DRY GAS EXPIRY DATE: 06/29/2019
VIOLATION CODE: DUI
ARREST TIME: 04:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: MCDANIEL

DATE/TIME OF TEST: 01/09/2017 04:22:18

Diagnostics Check	OK	04:22	
Air Blank		04:22	RFI Detect
Air Blank		04:22	RFI Detect

Data Download Date/Time: 01/26/2017 00:47

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003411

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANGELICCHI,JOSEPH,S
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: ANGELICCHI,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 477627
DRY GAS EXPIRY DATE: 06/29/2019
VIOLATION CODE: DUI
ARREST TIME: 22:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: KURAYA-ZIYADEH

DATE/TIME OF TEST: 01/09/2017 23:38:36

Diagnostics Check		OK	23:32
Air Blank		0.000	23:32
Control Test		0.081	23:33
Air Blank		0.000	23:33
Subject Sample	# 1	0.010	23:34
Breath Volume		4.832	23:34
Air Blank		0.000	23:34
Air Blank		0.000	23:36
Subject Sample	# 2	0.008	23:37
Breath Volume		2.414	23:37
Air Blank		0.000	23:37
Control Test		0.080	23:38
Air Blank		0.000	23:38
Diagnostics Check		OK	23:38

Data Download Date/Time: 01/26/2017 00:47

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003411

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANGELICCHI,JOSEPH,S
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: ANGELICCHI,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 477627
DRY GAS EXPIRY DATE: 06/29/2019
VIOLATION CODE: DUI
ARREST TIME: 02:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: LIGHT

DATE/TIME OF TEST: 01/21/2017 03:18:22

Diagnostics Check		OK	03:11
Air Blank		0.000	03:12
Control Test		0.078	03:12
Air Blank		0.000	03:13
Subject Sample	# 1	0.151	03:13
Breath Volume		4.000	03:13
Air Blank		0.000	03:14
Air Blank		0.000	03:16
Subject Sample	# 2	0.145	03:16
Breath Volume		3.523	03:16
Air Blank		0.000	03:17
Control Test		0.076	03:17
Air Blank		0.000	03:18
Diagnostics Check		OK	03:18

Data Download Date/Time: 01/26/2017 00:47

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003411

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANGELICCHI,JOSEPH,S
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: ANGELICCHI,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 477627
DRY GAS EXPIRY DATE: 06/29/2019
VIOLATION CODE: DUI
ARREST TIME: 23:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: COMFORT

DATE/TIME OF TEST: 01/22/2017 00:27:09

Diagnostics Check		OK	00:21
Air Blank		0.000	00:21
Control Test		0.079	00:21
Air Blank		0.000	00:22
Subject Sample	# 1	0.233	00:22
Breath Volume		2.234	00:22
Air Blank		0.000	00:23
Air Blank		0.000	00:25
Subject Sample	# 2	0.227	00:25
Breath Volume		2.613	00:25
Air Blank		0.000	00:26
Control Test		0.079	00:26
Air Blank		0.000	00:27
Diagnostics Check		OK	00:27

Data Download Date/Time: 01/26/2017 00:47

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003411

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPBELL,JOSHUA,C
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: CAMPBELL,JOSHUA
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 477627
DRY GAS EXPIRY DATE: 06/29/2019
VIOLATION CODE: DUI
ARREST TIME: 22:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:15

SUBJECT'S LAST NAME: HARVEY

DATE/TIME OF TEST: 01/28/2017 22:42:15

Diagnostics Check		OK	22:35
Air Blank		0.000	22:36
Control Test		0.079	22:36
Air Blank		0.000	22:37
Subject Sample	# 1	0.067	22:37
Breath Volume		2.039	22:37
Air Blank		0.000	22:38
Air Blank		0.000	22:40
Subject Sample	# 2	0.069	22:40
Breath Volume		3.019	22:40
Air Blank		0.000	22:41
Control Test		0.080	22:41
Air Blank		0.000	22:42
Diagnostics Check		OK	22:42

Data Download Date/Time: 02/27/2017 23:58

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003413

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNOWLES,CONSTANCE,J
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: KNOWLES,CONSTANCE
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 01/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A3
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:59

SUBJECT'S LAST NAME: DEAN

DATE/TIME OF TEST: 01/28/2017 02:25:38

Diagnostics Check		OK	02:19
Air Blank		0.000	02:19
Control Test		0.079	02:20
Air Blank		0.000	02:20
Subject Sample	# 1	0.302	02:21
Breath Volume		2.332	02:21
Air Blank		0.000	02:21
Air Blank		0.000	02:23
Subject Sample	# 2	0.300	02:24
Breath Volume		2.261	02:24
Air Blank		0.000	02:24
Control Test		0.079	02:25
Air Blank		0.000	02:25
Diagnostics Check		OK	02:25

Data Download Date/Time: 03/01/2017 15:30

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004587

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BEEKMAN,JACOB,C
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: BEEKMAN,JACOB
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 12/05/2016

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 02:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:47

SUBJECT'S LAST NAME: DIORIO

DATE/TIME OF TEST: 01/15/2017 02:17:27

Diagnostics Check		OK	02:14	
Air Blank		0.000	02:14	
Control Test		0.078	02:14	
Air Blank		0.000	02:15	
Subject Sample	# 1	0.071	02:16	
Breath Volume		1.996	02:16	
Air Blank			02:16	Ambient Fail
Air Blank		0.000	02:17	

Data Download Date/Time: 01/28/2017 15:32

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004587

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BEEKMAN,JACOB,C
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: BEEKMAN,JACOB
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 12/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 02:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:47

SUBJECT'S LAST NAME: DIORIO

DATE/TIME OF TEST: 01/15/2017 02:27:12

Diagnostics Check		OK	02:20
Air Blank		0.000	02:20
Control Test		0.078	02:21
Air Blank		0.000	02:21
Subject Sample	# 1	0.062	02:22
Breath Volume		1.476	02:22
Air Blank		0.000	02:22
Air Blank		0.000	02:24
Subject Sample	# 2	0.058	02:25
Breath Volume		1.679	02:25
Air Blank		0.000	02:26
Control Test		0.077	02:26
Air Blank		0.000	02:27
Diagnostics Check		OK	02:27

Data Download Date/Time: 01/28/2017 15:32

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SARMENT,JOSEPH,D
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: SARMENT,JOSEPH
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 12/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:37

SUBJECT'S LAST NAME: FOSTER

DATE/TIME OF TEST: 01/09/2017 01:10:15

Diagnostics Check		OK	01:07	
Air Blank		0.000	01:07	
Control Test		0.077	01:08	
Air Blank		0.000	01:08	
Subject Sample	# 1		01:08	Subject Test Refused
Breath Volume		0.000	01:08	
Air Blank		0.000	01:09	
Control Test		0.077	01:09	
Air Blank		0.000	01:10	
Diagnostics Check		OK	01:10	

Data Download Date/Time: 01/25/2017 14:29

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SARMENT,JOSEPH,D
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: SARMENT,JOSEPH
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 12/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:04

SUBJECT'S LAST NAME: FOSTER

DATE/TIME OF TEST: 01/09/2017 01:34:15

Diagnostics Check		OK	01:31	
Air Blank		0.000	01:31	
Control Test		0.078	01:32	
Air Blank		0.000	01:32	
Subject Sample	# 1		01:32	Subject Test Refused
Breath Volume		0.000	01:32	
Air Blank		0.000	01:33	
Control Test		0.078	01:33	
Air Blank		0.000	01:34	
Diagnostics Check		OK	01:34	

Data Download Date/Time: 01/25/2017 14:29

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,WILLIE,G
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: FREMBGEN,TIA
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 12/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:43

SUBJECT'S LAST NAME: TEEHAN

DATE/TIME OF TEST: 01/19/2017 00:10:34

Diagnostics Check		OK	00:03
Air Blank		0.000	00:04
Control Test		0.077	00:04
Air Blank		0.000	00:05
Subject Sample	# 1	0.064	00:05
Breath Volume		1.296	00:05
Air Blank		0.000	00:06
Air Blank		0.000	00:08
Subject Sample	# 2	0.062	00:09
Breath Volume		1.757	00:09
Air Blank		0.000	00:09
Control Test		0.076	00:09
Air Blank		0.000	00:10
Diagnostics Check		OK	00:10

Data Download Date/Time: 01/25/2017 14:29

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADKINS,JOSEPH,L
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: ADKINS,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 12/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 15:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:00

SUBJECT'S LAST NAME: MADDOX

DATE/TIME OF TEST: 01/20/2017 19:23:43

Diagnostics Check		OK	19:17
Air Blank		0.000	19:17
Control Test		0.076	19:18
Air Blank		0.000	19:18
Subject Sample	# 1	0.243	19:19
Breath Volume		2.359	19:19
Air Blank		0.000	19:19
Air Blank		0.000	19:21
Subject Sample	# 2	0.244	19:22
Breath Volume		2.062	19:22
Air Blank		0.000	19:22
Control Test		0.075	19:23
Air Blank		0.000	19:23
Diagnostics Check		OK	19:23

Data Download Date/Time: 01/25/2017 14:29

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005058

INSTRUMENT REGISTERED TO: CRESTVIEW PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SEALS,JOSEPH,J
BREATH TEST OPERATOR AGENCY: Crestview PD
ARRESTING OFFICER: SEALS,JOSEPH
ARRESTING OFFICER AGENCY: Crestview PD
LAST AGENCY INSPECTION: 12/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: SANDERS

DATE/TIME OF TEST: 01/01/2017 01:57:50

Diagnostics Check		OK	01:51
Air Blank		0.000	01:51
Control Test		0.078	01:52
Air Blank		0.000	01:52
Subject Sample	# 1	0.228	01:53
Breath Volume		2.996	01:53
Air Blank		0.000	01:54
Air Blank		0.000	01:55
Subject Sample	# 2	0.220	01:56
Breath Volume		1.933	01:56
Air Blank		0.000	01:56
Control Test		0.077	01:57
Air Blank		0.000	01:57
Diagnostics Check		OK	01:57

Data Download Date/Time: 01/31/2017 15:12

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005058

INSTRUMENT REGISTERED TO: CRESTVIEW PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAUN,DANIEL,J
BREATH TEST OPERATOR AGENCY: Crestview PD
ARRESTING OFFICER: NEWCOMB,COREY
ARRESTING OFFICER AGENCY: Crestview PD
LAST AGENCY INSPECTION: 12/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 16:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:00

SUBJECT'S LAST NAME: DAY

DATE/TIME OF TEST: 01/14/2017 17:30:08

Diagnostics Check		OK	17:21
Air Blank		0.000	17:21
Control Test		0.079	17:21
Air Blank		0.000	17:22
Subject Sample	# 1	0.075	17:23
Breath Volume		1.914	17:23
Air Blank		0.000	17:24
Air Blank		0.000	17:26
Subject Sample	# 2	0.072	17:28
Breath Volume		2.242	17:28
Air Blank		0.000	17:29
Control Test		0.078	17:29
Air Blank		0.000	17:30
Diagnostics Check		OK	17:30

Data Download Date/Time: 01/31/2017 15:12

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005076

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FOWLER,TYLER,J
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: FOWLER,TYLER
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 12/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:43

SUBJECT'S LAST NAME: HOLAN

DATE/TIME OF TEST: 01/01/2017 21:19:17

Diagnostics Check		OK	21:11
Air Blank		0.000	21:11
Control Test		0.078	21:12
Air Blank		0.000	21:12
Subject Sample	# 1	0.055	21:14
Breath Volume		1.878	21:14
Air Blank		0.000	21:15
Air Blank		0.000	21:17
Subject Sample	# 2	0.050	21:17
Breath Volume		2.039	21:17
Air Blank		0.000	21:18
Control Test		0.077	21:18
Air Blank		0.000	21:19
Diagnostics Check		OK	21:19

Data Download Date/Time: 01/27/2017 08:54

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005076

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FOWLER,TYLER,J
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: FOWLER,TYLER
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 12/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:18

SUBJECT'S LAST NAME: MONKS

DATE/TIME OF TEST: 01/11/2017 21:44:33

Diagnostics Check		OK	21:38
Air Blank		0.000	21:38
Control Test		0.079	21:39
Air Blank		0.000	21:39
Subject Sample	# 1	0.126	21:40
Breath Volume		4.847	21:40
Air Blank		0.000	21:40
Air Blank		0.000	21:42
Subject Sample	# 2	0.126	21:42
Breath Volume		4.785	21:42
Air Blank		0.000	21:43
Control Test		0.077	21:43
Air Blank		0.000	21:44
Diagnostics Check		OK	21:44

Data Download Date/Time: 01/27/2017 08:54

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005076

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FOWLER,TYLER,J
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: FOWLER,TYLER
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 12/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: MORROW

DATE/TIME OF TEST: 01/14/2017 02:09:10

Diagnostics Check		OK	02:05	
Air Blank		0.000	02:06	
Control Test		0.079	02:06	
Air Blank		0.000	02:07	
Subject Sample	# 1		02:07	Subject Test Refused
Breath Volume		0.000	02:07	
Air Blank		0.000	02:08	
Control Test		0.078	02:08	
Air Blank		0.000	02:09	
Diagnostics Check		OK	02:09	

Data Download Date/Time: 01/27/2017 08:54

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005076

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FOWLER,TYLER,J
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: FOWLER,TYLER
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 12/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:07

SUBJECT'S LAST NAME: PENTSA

DATE/TIME OF TEST: 01/17/2017 21:35:34

Diagnostics Check		OK	21:28
Air Blank		0.000	21:28
Control Test		0.079	21:29
Air Blank		0.000	21:29
Subject Sample	# 1	0.082	21:30
Breath Volume		2.386	21:30
Air Blank		0.000	21:31
Air Blank		0.000	21:33
Subject Sample	# 2	0.088	21:34
Breath Volume		3.445	21:34
Air Blank		0.000	21:34
Control Test		0.078	21:34
Air Blank		0.000	21:35
Diagnostics Check		OK	21:35

Data Download Date/Time: 01/27/2017 08:54

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005076

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FOWLER,TYLER,J
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: FOWLER,TYLER
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 12/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:17

SUBJECT'S LAST NAME: KRASKO

DATE/TIME OF TEST: 01/19/2017 23:44:04

Diagnostics Check		OK	23:37
Air Blank		0.000	23:38
Control Test		0.079	23:38
Air Blank		0.000	23:38
Subject Sample	# 1	0.102	23:39
Breath Volume		1.851	23:39
Air Blank		0.000	23:39
Air Blank		0.000	23:41
Subject Sample	# 2	0.109	23:42
Breath Volume		1.992	23:42
Air Blank		0.000	23:43
Control Test		0.077	23:43
Air Blank		0.000	23:43
Diagnostics Check		OK	23:44

Data Download Date/Time: 01/27/2017 08:54

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005076

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FOWLER],TYLER,J
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: FOWLER,TYLER
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 12/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:38

SUBJECT'S LAST NAME: DOZIER

DATE/TIME OF TEST: 01/25/2017 22:15:20

Diagnostics Check		OK	22:08
Air Blank		0.000	22:08
Control Test		0.078	22:08
Air Blank		0.000	22:09
Subject Sample	# 1	0.083	22:10
Breath Volume		1.562	22:10
Air Blank		0.000	22:11
Air Blank		0.000	22:13
Subject Sample	# 2	0.077	22:13
Breath Volume		1.152	22:13
Air Blank		0.000	22:14
Control Test		0.077	22:14
Air Blank		0.000	22:15
Diagnostics Check		OK	22:15

Data Download Date/Time: 01/27/2017 08:54

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005077

INSTRUMENT REGISTERED TO: NORTH MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PINEDA,MIGUEL,A
BREATH TEST OPERATOR AGENCY: North Miami PD
ARRESTING OFFICER: FERNANDEZ,Y
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 01/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06415080A2
DRY GAS EXPIRY DATE: 04/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:22

SUBJECT'S LAST NAME: CAPORAL

DATE/TIME OF TEST: 01/22/2017 01:52:32

Diagnostics Check		OK	01:49	
Air Blank		0.000	01:50	
Control Test		0.079	01:50	
Air Blank		0.000	01:50	
Subject Sample	# 1		01:51	Subject Test Refused
Breath Volume		0.000	01:51	
Air Blank		0.000	01:51	
Control Test		0.078	01:51	
Air Blank		0.000	01:52	
Diagnostics Check		OK	01:52	

Data Download Date/Time: 02/14/2017 11:43

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005077

INSTRUMENT REGISTERED TO: NORTH MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PINEDA,MIGUEL,A
BREATH TEST OPERATOR AGENCY: North Miami PD
ARRESTING OFFICER: PINEDA,MIGUEL
ARRESTING OFFICER AGENCY: North Miami PD
LAST AGENCY INSPECTION: 01/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06415080A2
DRY GAS EXPIRY DATE: 04/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:36

SUBJECT'S LAST NAME: BEAUVILLE

DATE/TIME OF TEST: 01/23/2017 03:05:43

Diagnostics Check		OK	02:59
Air Blank		0.000	03:00
Control Test		0.078	03:00
Air Blank		0.000	03:00
Subject Sample	# 1	0.121	03:01
Breath Volume		2.632	03:01
Air Blank		0.000	03:01
Air Blank		0.000	03:03
Subject Sample	# 2	0.124	03:04
Breath Volume		2.710	03:04
Air Blank		0.000	03:04
Control Test		0.078	03:05
Air Blank		0.000	03:05
Diagnostics Check		OK	03:05

Data Download Date/Time: 02/14/2017 11:43

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005077

INSTRUMENT REGISTERED TO: NORTH MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PINEDA,MIGUEL,A
BREATH TEST OPERATOR AGENCY: North Miami PD
ARRESTING OFFICER: MONTALVAN,LUIS
ARRESTING OFFICER AGENCY: North Miami PD
LAST AGENCY INSPECTION: 01/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06415080A2
DRY GAS EXPIRY DATE: 04/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:38

SUBJECT'S LAST NAME: TRISTA

DATE/TIME OF TEST: 01/23/2017 04:42:20

Diagnostics Check		OK	04:34
Air Blank		0.000	04:35
Control Test		0.078	04:35
Air Blank		0.000	04:35
Subject Sample	# 1	0.061	04:37
Breath Volume		2.949	04:37
Air Blank		0.000	04:38
Air Blank		0.000	04:40
Subject Sample	# 2	0.054	04:40
Breath Volume		1.171	04:40
Air Blank		0.000	04:41
Control Test		0.078	04:41
Air Blank		0.000	04:42
Diagnostics Check		OK	04:42

Data Download Date/Time: 02/14/2017 11:43

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRENNAN,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:41

SUBJECT'S LAST NAME: MABRY II

DATE/TIME OF TEST: 01/10/2017 02:19:16

Diagnostics Check		OK	02:12
Air Blank		0.000	02:13
Control Test		0.079	02:13
Air Blank		0.000	02:14
Subject Sample	# 1	0.164	02:14
Breath Volume		3.316	02:14
Air Blank		0.000	02:15
Air Blank		0.000	02:17
Subject Sample	# 2	0.163	02:17
Breath Volume		3.113	02:17
Air Blank		0.000	02:18
Control Test		0.080	02:18
Air Blank		0.000	02:19
Diagnostics Check		OK	02:19

Data Download Date/Time: 01/10/2017 18:29

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PECYLAK,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: KINDELL

DATE/TIME OF TEST: 01/10/2017 04:23:13

Diagnostics Check		OK	04:13	
Air Blank		0.000	04:13	
Control Test		0.080	04:14	
Air Blank		0.000	04:14	
Subject Sample	# 1		04:17	No Sample Provided
Breath Volume		0.000	04:17	
Air Blank		0.000	04:18	
Air Blank		0.000	04:20	
Subject Sample	# 2		04:21	Subject Test Refused
Breath Volume		0.000	04:21	
Air Blank		0.000	04:22	
Control Test		0.079	04:22	
Air Blank		0.000	04:23	
Diagnostics Check		OK	04:23	

Data Download Date/Time: 01/10/2017 18:29

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:06

SUBJECT'S LAST NAME: WAN

DATE/TIME OF TEST: 01/12/2017 04:32:34

Diagnostics Check		OK	04:26
Air Blank		0.000	04:26
Control Test		0.079	04:27
Air Blank		0.000	04:27
Subject Sample	# 1	0.100	04:28
Breath Volume		2.968	04:28
Air Blank		0.000	04:28
Air Blank		0.000	04:30
Subject Sample	# 2	0.097	04:31
Breath Volume		2.558	04:31
Air Blank		0.000	04:31
Control Test		0.079	04:31
Air Blank		0.000	04:32
Diagnostics Check		OK	04:32

Data Download Date/Time: 02/08/2017 17:44

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRIGHTWELL,R
ARRESTING OFFICER AGENCY: U of So FL Campus PD
LAST AGENCY INSPECTION: 01/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:54

SUBJECT'S LAST NAME: ALSHEHRI

DATE/TIME OF TEST: 01/13/2017 02:21:12

Diagnostics Check		OK	02:14
Air Blank		0.000	02:15
Control Test		0.079	02:15
Air Blank		0.000	02:15
Subject Sample	# 1	0.073	02:16
Breath Volume		2.785	02:16
Air Blank		0.000	02:17
Air Blank		0.000	02:19
Subject Sample	# 2	0.073	02:19
Breath Volume		2.886	02:19
Air Blank		0.000	02:20
Control Test		0.079	02:20
Air Blank		0.000	02:21
Diagnostics Check		OK	02:21

Data Download Date/Time: 02/08/2017 17:44

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRENNAN,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:31

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 01/13/2017 02:59:19

Diagnostics Check		OK	02:52
Air Blank		0.000	02:53
Control Test		0.079	02:53
Air Blank		0.000	02:54
Subject Sample	# 1	0.173	02:54
Breath Volume		2.843	02:54
Air Blank		0.000	02:55
Air Blank		0.000	02:57
Subject Sample	# 2	0.172	02:57
Breath Volume		2.292	02:57
Air Blank		0.000	02:58
Control Test		0.079	02:58
Air Blank		0.000	02:59
Diagnostics Check		OK	02:59

Data Download Date/Time: 02/08/2017 17:44

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY,PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: EDGE,S
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:32

SUBJECT'S LAST NAME: BALDERAS

DATE/TIME OF TEST: 01/13/2017 21:02:02

Diagnostics Check		OK	20:52	
Air Blank		0.000	20:53	
Control Test		0.080	20:53	
Air Blank		0.000	20:53	
Subject Sample	# 1	0.000	20:57	Volume Not Met
Breath Volume		0.082	20:57	
Air Blank		0.000	20:57	
Air Blank		0.000	20:59	
Subject Sample	# 2		21:00	Subject Test Refused
Breath Volume		0.015	21:00	
Air Blank		0.000	21:01	
Control Test		0.082	21:01	
Air Blank		0.000	21:01	
Diagnostics Check		OK	21:02	

Data Download Date/Time: 02/08/2017 17:44

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:14

SUBJECT'S LAST NAME: QUINTANA

DATE/TIME OF TEST: 01/14/2017 04:42:01

Diagnostics Check		OK	04:34
Air Blank		0.000	04:34
Control Test		0.079	04:35
Air Blank		0.000	04:35
Subject Sample	# 1	0.063	04:36
Breath Volume		1.218	04:36
Air Blank		0.000	04:37
Air Blank		0.000	04:39
Subject Sample	# 2	0.056	04:40
Breath Volume		1.164	04:40
Air Blank		0.000	04:41
Control Test		0.081	04:41
Air Blank		0.000	04:41
Diagnostics Check		OK	04:42

Data Download Date/Time: 02/08/2017 17:44

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY,PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GONZALEZ,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:37

SUBJECT'S LAST NAME: JANEK

DATE/TIME OF TEST: 01/14/2017 05:05:04

Diagnostics Check		OK	04:58
Air Blank		0.000	04:58
Control Test		0.079	04:58
Air Blank		0.000	04:59
Subject Sample	# 1	0.116	05:00
Breath Volume		3.007	05:00
Air Blank		0.000	05:01
Air Blank		0.000	05:02
Subject Sample	# 2	0.117	05:03
Breath Volume		2.324	05:03
Air Blank		0.000	05:04
Control Test		0.080	05:04
Air Blank		0.000	05:04
Diagnostics Check		OK	05:05

Data Download Date/Time: 02/08/2017 17:44

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:33

SUBJECT'S LAST NAME: HART

DATE/TIME OF TEST: 01/15/2017 01:00:13

Diagnostics Check		OK	00:53
Air Blank		0.000	00:54
Control Test		0.080	00:54
Air Blank		0.000	00:54
Subject Sample	# 1	0.172	00:55
Breath Volume		2.406	00:55
Air Blank		0.000	00:56
Air Blank		0.000	00:58
Subject Sample	# 2	0.192	00:58
Breath Volume		2.847	00:58
Air Blank		0.000	00:59
Control Test		0.079	00:59
Air Blank		0.000	01:00
Diagnostics Check		OK	01:00

Data Download Date/Time: 02/08/2017 17:44

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WOODEN,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: FITZGERALD

DATE/TIME OF TEST: 01/16/2017 01:52:21

Diagnostics Check		OK	01:45
Air Blank		0.000	01:45
Control Test		0.080	01:46
Air Blank		0.000	01:46
Subject Sample	# 1	0.175	01:47
Breath Volume		1.816	01:47
Air Blank		0.000	01:48
Air Blank		0.000	01:50
Subject Sample	# 2	0.165	01:50
Breath Volume		1.187	01:50
Air Blank		0.000	01:51
Control Test		0.079	01:51
Air Blank		0.000	01:52
Diagnostics Check		OK	01:52

Data Download Date/Time: 02/08/2017 17:44

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARLAUG,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: TRUJILLO

DATE/TIME OF TEST: 01/17/2017 01:28:05

Diagnostics Check		OK	01:20
Air Blank		0.000	01:20
Control Test		0.079	01:21
Air Blank		0.000	01:21
Subject Sample	# 1	0.125	01:23
Breath Volume		1.683	01:23
Air Blank		0.000	01:23
Air Blank		0.000	01:25
Subject Sample	# 2	0.133	01:26
Breath Volume		3.046	01:26
Air Blank		0.000	01:27
Control Test		0.080	01:27
Air Blank		0.000	01:28
Diagnostics Check		OK	01:28

Data Download Date/Time: 02/08/2017 17:44

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VAN-TREESE,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: FARAH

DATE/TIME OF TEST: 01/17/2017 02:41:51

Diagnostics Check		OK	02:30	
Air Blank		0.000	02:30	
Control Test		0.079	02:31	
Air Blank		0.000	02:31	
Subject Sample	# 1	0.043	02:34	Volume Not Met
Breath Volume		0.015	02:34	
Air Blank		0.000	02:35	
Air Blank		0.000	02:37	
Subject Sample	# 2	0.042	02:40	Volume Not Met
Breath Volume		0.066	02:40	
Air Blank		0.000	02:40	
Control Test		0.081	02:41	
Air Blank		0.000	02:41	
Diagnostics Check		OK	02:41	

Data Download Date/Time: 02/08/2017 17:44

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: EHRHART,M
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:57

SUBJECT'S LAST NAME: ARIAS CENTENO

DATE/TIME OF TEST: 01/18/2017 04:26:52

Diagnostics Check		OK	04:20
Air Blank		0.000	04:20
Control Test		0.079	04:20
Air Blank		0.000	04:21
Subject Sample	# 1	0.231	04:22
Breath Volume		3.835	04:22
Air Blank		0.000	04:23
Air Blank		0.000	04:24
Subject Sample	# 2	0.246	04:25
Breath Volume		3.039	04:25
Air Blank		0.000	04:25
Control Test		0.079	04:26
Air Blank		0.000	04:26
Diagnostics Check		OK	04:26

Data Download Date/Time: 02/08/2017 17:44

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY,PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GONZALEZ,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:54

SUBJECT'S LAST NAME: MICHAEL

DATE/TIME OF TEST: 01/19/2017 04:26:01

Diagnostics Check		OK	04:19
Air Blank		0.000	04:19
Control Test		0.080	04:19
Air Blank		0.000	04:20
Subject Sample	# 1	0.145	04:21
Breath Volume		1.332	04:21
Air Blank		0.000	04:22
Air Blank		0.000	04:24
Subject Sample	# 2	0.149	04:24
Breath Volume		1.425	04:24
Air Blank		0.000	04:25
Control Test		0.079	04:25
Air Blank		0.000	04:25
Diagnostics Check		OK	04:26

Data Download Date/Time: 02/08/2017 17:44

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ANDERSON,C
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:08

SUBJECT'S LAST NAME: POWELL

DATE/TIME OF TEST: 01/20/2017 05:40:54

Diagnostics Check		OK	05:29	
Air Blank		0.000	05:29	
Control Test		0.079	05:29	
Air Blank		0.000	05:30	
Subject Sample	# 1	0.180	05:33	Volume Not Met
Breath Volume		0.121	05:33	
Air Blank		0.000	05:34	
Air Blank		0.000	05:35	
Subject Sample	# 2	0.194	05:39	Volume Not Met
Breath Volume		0.113	05:39	
Air Blank		0.000	05:39	
Control Test		0.081	05:40	
Air Blank		0.000	05:40	
Diagnostics Check		OK	05:40	

Data Download Date/Time: 02/08/2017 17:44

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ANDERSON,C
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:08

SUBJECT'S LAST NAME: POWELL

DATE/TIME OF TEST: 01/20/2017 05:47:28

Diagnostics Check		OK	05:44	
Air Blank		0.000	05:45	
Control Test		0.079	05:45	
Air Blank		0.000	05:45	
Subject Sample	# 1		05:46	Subject Test Refused
Breath Volume		0.000	05:46	
Air Blank		0.000	05:46	
Control Test		0.078	05:46	
Air Blank		0.000	05:47	
Diagnostics Check		OK	05:47	

Data Download Date/Time: 02/08/2017 17:44

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DAWSY,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:44

SUBJECT'S LAST NAME: SANTIAGO

DATE/TIME OF TEST: 01/21/2017 02:11:52

Diagnostics Check		OK	02:04
Air Blank		0.000	02:05
Control Test		0.079	02:05
Air Blank		0.000	02:06
Subject Sample	# 1	0.100	02:07
Breath Volume		2.597	02:07
Air Blank		0.000	02:07
Air Blank		0.000	02:09
Subject Sample	# 2	0.099	02:10
Breath Volume		2.562	02:10
Air Blank		0.000	02:10
Control Test		0.079	02:11
Air Blank		0.000	02:11
Diagnostics Check		OK	02:11

Data Download Date/Time: 02/08/2017 17:44

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARLAUG,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:58

SUBJECT'S LAST NAME: BRADLEY

DATE/TIME OF TEST: 01/21/2017 04:25:14

Diagnostics Check		OK	04:18
Air Blank		0.000	04:19
Control Test		0.079	04:19
Air Blank		0.000	04:19
Subject Sample	# 1	0.190	04:20
Breath Volume		1.972	04:20
Air Blank		0.000	04:21
Air Blank		0.000	04:23
Subject Sample	# 2	0.189	04:23
Breath Volume		1.937	04:23
Air Blank		0.000	04:24
Control Test		0.079	04:24
Air Blank		0.000	04:25
Diagnostics Check		OK	04:25

Data Download Date/Time: 02/08/2017 17:44

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY,PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FRANKLIN,R
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:09

SUBJECT'S LAST NAME: SILVERA

DATE/TIME OF TEST: 01/22/2017 01:43:34

Diagnostics Check		OK	01:34
Air Blank		0.000	01:35
Control Test		0.080	01:35
Air Blank		0.000	01:36
Subject Sample	# 1	0.159	01:38
Breath Volume		1.957	01:38
Air Blank		0.000	01:39
Air Blank		0.000	01:41
Subject Sample	# 2	0.160	01:41
Breath Volume		1.632	01:41
Air Blank		0.000	01:42
Control Test		0.080	01:42
Air Blank		0.000	01:43
Diagnostics Check		OK	01:43

Data Download Date/Time: 02/08/2017 17:44

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: NIEUWENDAM,R
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:43

SUBJECT'S LAST NAME: MOYE

DATE/TIME OF TEST: 01/22/2017 03:10:09

Diagnostics Check		OK	03:03
Air Blank		0.000	03:04
Control Test		0.080	03:04
Air Blank		0.000	03:05
Subject Sample	# 1	0.132	03:05
Breath Volume		2.292	03:05
Air Blank		0.000	03:06
Air Blank		0.000	03:08
Subject Sample	# 2	0.131	03:08
Breath Volume		2.531	03:08
Air Blank		0.000	03:09
Control Test		0.080	03:09
Air Blank		0.000	03:10
Diagnostics Check		OK	03:10

Data Download Date/Time: 02/08/2017 17:44

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BLANCHARD,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 01/22/2017 03:51:01

Diagnostics Check		OK	03:44
Air Blank		0.000	03:44
Control Test		0.079	03:45
Air Blank		0.000	03:45
Subject Sample	# 1	0.120	03:46
Breath Volume		2.136	03:46
Air Blank		0.000	03:46
Air Blank		0.000	03:48
Subject Sample	# 2	0.130	03:49
Breath Volume		2.582	03:49
Air Blank		0.000	03:50
Control Test		0.079	03:50
Air Blank		0.000	03:50
Diagnostics Check		OK	03:51

Data Download Date/Time: 02/08/2017 17:44

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARLAUG,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:27

SUBJECT'S LAST NAME: ROUBALOVA

DATE/TIME OF TEST: 01/26/2017 03:56:05

Diagnostics Check		OK	03:47
Air Blank		0.000	03:47
Control Test		0.079	03:48
Air Blank		0.000	03:48
Subject Sample	# 1	0.176	03:51
Breath Volume		1.113	03:51
Air Blank		0.000	03:51
Air Blank		0.000	03:53
Subject Sample	# 2	0.171	03:54
Breath Volume		1.632	03:54
Air Blank		0.000	03:55
Control Test		0.080	03:55
Air Blank		0.000	03:55
Diagnostics Check		OK	03:56

Data Download Date/Time: 02/08/2017 17:44

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DAWSY,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 01/27/2017 02:12:16

Diagnostics Check		OK	02:05
Air Blank		0.000	02:06
Control Test		0.080	02:06
Air Blank		0.000	02:07
Subject Sample	# 1	0.151	02:07
Breath Volume		2.796	02:07
Air Blank		0.000	02:08
Air Blank		0.000	02:10
Subject Sample	# 2	0.135	02:10
Breath Volume		2.300	02:10
Air Blank		0.000	02:11
Control Test		0.079	02:11
Air Blank		0.000	02:12
Diagnostics Check		OK	02:12

Data Download Date/Time: 02/08/2017 17:45

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DAWSY,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:58

SUBJECT'S LAST NAME: HERNANDEZ CASTRO

DATE/TIME OF TEST: 01/28/2017 20:25:03

Diagnostics Check		OK	20:18
Air Blank		0.000	20:18
Control Test		0.080	20:19
Air Blank		0.000	20:19
Subject Sample	# 1	0.202	20:20
Breath Volume		1.421	20:20
Air Blank		0.000	20:21
Air Blank		0.000	20:22
Subject Sample	# 2	0.210	20:23
Breath Volume		2.261	20:23
Air Blank		0.000	20:24
Control Test		0.080	20:24
Air Blank		0.000	20:24
Diagnostics Check		OK	20:25

Data Download Date/Time: 02/08/2017 17:45

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GONZALEZ,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:02

SUBJECT'S LAST NAME: HILL

DATE/TIME OF TEST: 01/28/2017 21:28:52

Diagnostics Check		OK	21:22
Air Blank		0.000	21:22
Control Test		0.079	21:23
Air Blank		0.000	21:23
Subject Sample	# 1	0.099	21:24
Breath Volume		2.312	21:24
Air Blank		0.000	21:25
Air Blank		0.000	21:26
Subject Sample	# 2	0.099	21:27
Breath Volume		2.878	21:27
Air Blank		0.000	21:27
Control Test		0.080	21:28
Air Blank		0.000	21:28
Diagnostics Check		OK	21:28

Data Download Date/Time: 02/08/2017 17:45

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COCHRANE,AARON,M
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: COCHRANE,AARON
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 12/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:44

SUBJECT'S LAST NAME: MCNULTY

DATE/TIME OF TEST: 01/01/2017 01:11:12

Diagnostics Check		OK	01:04
Air Blank		0.000	01:05
Control Test		0.082	01:05
Air Blank		0.000	01:06
Subject Sample	# 1	0.000	01:06
Breath Volume		2.679	01:06
Air Blank		0.000	01:07
Air Blank		0.000	01:09
Subject Sample	# 2	0.000	01:09
Breath Volume		2.535	01:09
Air Blank		0.000	01:10
Control Test		0.080	01:10
Air Blank		0.000	01:11
Diagnostics Check		OK	01:11

Data Download Date/Time: 01/19/2017 18:38

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVENPORT,JOSHUA,J
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: DAVENPORT,JOSHUA
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: CRUZ

DATE/TIME OF TEST: 01/07/2017 04:19:14

Diagnostics Check		OK	04:07	
Air Blank		0.000	04:07	
Control Test		0.081	04:07	
Air Blank		0.000	04:08	
Subject Sample	# 1	0.183	04:11	Volume Not Met
Breath Volume		0.976	04:11	
Air Blank		0.000	04:12	
Air Blank		0.000	04:14	
Subject Sample	# 2	0.217	04:14	
Breath Volume		1.550	04:14	
Air Blank		0.000	04:15	
Air Blank		0.000	04:17	
Subject Sample	# 3	0.220	04:17	
Breath Volume		1.589	04:17	
Air Blank		0.000	04:18	
Control Test		0.080	04:18	
Air Blank		0.000	04:19	
Diagnostics Check		OK	04:19	

Data Download Date/Time: 01/19/2017 18:38

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WITHERELL,BENJAMIN,T
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: REID,W
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:12

SUBJECT'S LAST NAME: GRIFFITH

DATE/TIME OF TEST: 01/08/2017 01:35:24

Diagnostics Check		OK	01:32	
Air Blank		0.000	01:32	
Control Test		0.083	01:33	
Air Blank		0.000	01:33	
Subject Sample	# 1		01:34	Subject Test Refused
Breath Volume		0.000	01:34	
Air Blank		0.000	01:34	
Control Test		0.082	01:34	
Air Blank		0.000	01:35	
Diagnostics Check		OK	01:35	

Data Download Date/Time: 01/19/2017 18:38

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WITHERELL,BENJAMIN,T
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: WITHERELL,BENJAMIN
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 12/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:48

SUBJECT'S LAST NAME: MARVILL

DATE/TIME OF TEST: 01/08/2017 02:14:51

Diagnostics Check		OK	02:08
Air Blank		0.000	02:08
Control Test		0.083	02:09
Air Blank		0.000	02:09
Subject Sample	# 1	0.140	02:10
Breath Volume		2.089	02:10
Air Blank		0.000	02:10
Air Blank		0.000	02:12
Subject Sample	# 2	0.140	02:13
Breath Volume		2.058	02:13
Air Blank		0.000	02:13
Control Test		0.082	02:14
Air Blank		0.000	02:14
Diagnostics Check		OK	02:14

Data Download Date/Time: 01/19/2017 18:39

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVENPORT,JOSHUA,J
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: DAVENPORT,JOSHUA
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: VICKERS

DATE/TIME OF TEST: 01/09/2017 02:01:37

Diagnostics Check		OK	01:54
Air Blank		0.000	01:55
Control Test		0.084	01:55
Air Blank		0.000	01:56
Subject Sample	# 1	0.180	01:57
Breath Volume		1.304	01:57
Air Blank		0.000	01:57
Air Blank		0.000	01:59
Subject Sample	# 2	0.173	02:00
Breath Volume		1.640	02:00
Air Blank		0.000	02:00
Control Test		0.083	02:01
Air Blank		0.000	02:01
Diagnostics Check		OK	02:01

Data Download Date/Time: 01/19/2017 18:39

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WITHERELL,BENJAMIN,T
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: WITHERELL,BENJAMIN
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 12/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: WEEKS

DATE/TIME OF TEST: 01/09/2017 03:57:59

Diagnostics Check		OK	03:51
Air Blank		0.000	03:51
Control Test		0.082	03:52
Air Blank		0.000	03:52
Subject Sample	# 1	0.174	03:53
Breath Volume		3.285	03:53
Air Blank		0.000	03:54
Air Blank		0.000	03:55
Subject Sample	# 2	0.174	03:56
Breath Volume		2.765	03:56
Air Blank		0.000	03:57
Control Test		0.081	03:57
Air Blank		0.000	03:57
Diagnostics Check		OK	03:57

Data Download Date/Time: 01/19/2017 18:39

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COCHRANE,AARON,M
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: COCHRANE,AARON
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 12/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: SHERBURN

DATE/TIME OF TEST: 01/10/2017 00:42:24

Diagnostics Check		OK	00:35
Air Blank		0.000	00:36
Control Test		0.082	00:36
Air Blank		0.000	00:37
Subject Sample	# 1	0.180	00:37
Breath Volume		3.917	00:37
Air Blank		0.000	00:38
Air Blank		0.000	00:40
Subject Sample	# 2	0.173	00:40
Breath Volume		3.113	00:40
Air Blank		0.000	00:41
Control Test		0.080	00:41
Air Blank		0.000	00:42
Diagnostics Check		OK	00:42

Data Download Date/Time: 01/19/2017 18:39

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO,I,
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: HYLER,T
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 12/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:21

SUBJECT'S LAST NAME: HERZOG

DATE/TIME OF TEST: 01/10/2017 01:48:12

Diagnostics Check		OK	01:41
Air Blank		0.000	01:41
Control Test		0.082	01:42
Air Blank		0.000	01:42
Subject Sample	# 1	0.129	01:43
Breath Volume		2.332	01:43
Air Blank		0.000	01:43
Air Blank		0.000	01:45
Subject Sample	# 2	0.129	01:46
Breath Volume		1.906	01:46
Air Blank		0.000	01:47
Control Test		0.082	01:47
Air Blank		0.000	01:48
Diagnostics Check		OK	01:48

Data Download Date/Time: 01/19/2017 18:39

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRUZ,CHRISTINA,M
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: KENT,NATHAN
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 12/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 07:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:08

SUBJECT'S LAST NAME: CONNOLLY

DATE/TIME OF TEST: 01/13/2017 08:34:58

Diagnostics Check		OK	08:28
Air Blank		0.000	08:29
Control Test		0.082	08:29
Air Blank		0.000	08:29
Subject Sample	# 1	0.159	08:30
Breath Volume		2.828	08:30
Air Blank		0.000	08:31
Air Blank		0.000	08:33
Subject Sample	# 2	0.161	08:33
Breath Volume		2.925	08:33
Air Blank		0.000	08:34
Control Test		0.082	08:34
Air Blank		0.000	08:34
Diagnostics Check		OK	08:34

Data Download Date/Time: 01/19/2017 18:39

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO,I,
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: LAVELLE,D
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: MITCHELL

DATE/TIME OF TEST: 01/14/2017 01:16:46

Diagnostics Check		OK	01:10
Air Blank		0.000	01:10
Control Test		0.081	01:11
Air Blank		0.000	01:11
Subject Sample	# 1	0.175	01:12
Breath Volume		3.417	01:12
Air Blank		0.000	01:12
Air Blank		0.000	01:14
Subject Sample	# 2	0.161	01:15
Breath Volume		1.539	01:15
Air Blank		0.000	01:15
Control Test		0.081	01:16
Air Blank		0.000	01:16
Diagnostics Check		OK	01:16

Data Download Date/Time: 01/19/2017 18:39

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAOLELLA,MICHAEL,P
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: MAYO,R
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: HOWELL

DATE/TIME OF TEST: 01/15/2017 01:51:37

Diagnostics Check		OK	01:45
Air Blank		0.000	01:45
Control Test		0.082	01:45
Air Blank		0.000	01:46
Subject Sample	# 1	0.090	01:47
Breath Volume		3.050	01:47
Air Blank		0.000	01:47
Air Blank		0.000	01:49
Subject Sample	# 2	0.079	01:50
Breath Volume		1.339	01:50
Air Blank		0.000	01:50
Control Test		0.081	01:51
Air Blank		0.000	01:51
Diagnostics Check		OK	01:51

Data Download Date/Time: 01/19/2017 18:39

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COCHRANE,AARON,M
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: COCHRANE,AARON
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 12/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: MULLAHEY

DATE/TIME OF TEST: 01/16/2017 00:57:31

Diagnostics Check		OK	00:50
Air Blank		0.000	00:50
Control Test		0.081	00:50
Air Blank		0.000	00:51
Subject Sample	# 1	0.216	00:52
Breath Volume		1.812	00:52
Air Blank		0.000	00:53
Air Blank		0.000	00:54
Subject Sample	# 2	0.217	00:55
Breath Volume		2.343	00:55
Air Blank		0.000	00:56
Control Test		0.081	00:56
Air Blank		0.000	00:57
Diagnostics Check		OK	00:57

Data Download Date/Time: 01/19/2017 18:39

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WITHERELL,BENJAMIN,T
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: WITHERELL,BENJAMIN
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 12/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:53

SUBJECT'S LAST NAME: HALLEY

DATE/TIME OF TEST: 01/16/2017 22:19:58

Diagnostics Check		OK	22:13
Air Blank		0.000	22:14
Control Test		0.082	22:14
Air Blank		0.000	22:15
Subject Sample	# 1	0.085	22:15
Breath Volume		2.843	22:15
Air Blank		0.000	22:16
Air Blank		0.000	22:18
Subject Sample	# 2	0.081	22:18
Breath Volume		2.597	22:18
Air Blank		0.000	22:19
Control Test		0.080	22:19
Air Blank		0.000	22:19
Diagnostics Check		OK	22:19

Data Download Date/Time: 01/19/2017 18:39

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WITHERELL,BENJAMIN,T
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: WITHERELL,BENJAMIN
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 12/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: SANCHEZ

DATE/TIME OF TEST: 01/18/2017 02:23:24

Diagnostics Check		OK	02:10	
Air Blank		0.000	02:11	
Control Test		0.082	02:11	
Air Blank		0.000	02:11	
Subject Sample	# 1	0.195	02:13	
Breath Volume		1.515	02:13	
Air Blank		0.000	02:13	
Air Blank		0.000	02:15	
Subject Sample	# 2	0.166	02:18	Volume Not Met
Breath Volume		0.332	02:18	
Air Blank		0.000	02:19	
Air Blank		0.000	02:21	
Subject Sample	# 3	0.181	02:21	
Breath Volume		1.527	02:21	
Air Blank		0.000	02:22	
Control Test		0.077	02:22	
Air Blank		0.000	02:23	
Diagnostics Check		OK	02:23	

Data Download Date/Time: 01/19/2017 18:39

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOWATT,NATHAN,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MOWATT,NATHAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG622905
DRY GAS EXPIRY DATE: 08/16/2018
VIOLATION CODE: DUI
ARREST TIME: 01:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: FRIZZELL

DATE/TIME OF TEST: 01/01/2017 02:17:32

Diagnostics Check		OK	02:11
Air Blank		0.000	02:11
Control Test		0.082	02:11
Air Blank		0.000	02:12
Subject Sample	# 1	0.070	02:13
Breath Volume		1.964	02:13
Air Blank		0.000	02:13
Air Blank		0.000	02:15
Subject Sample	# 2	0.068	02:16
Breath Volume		1.511	02:16
Air Blank		0.000	02:16
Control Test		0.080	02:16
Air Blank		0.000	02:17
Diagnostics Check		OK	02:17

Data Download Date/Time: 01/12/2017 23:36

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POLANCO,KEDWIN,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: POLANCO,KEDWIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG622905
DRY GAS EXPIRY DATE: 08/16/2018
VIOLATION CODE: DUI
ARREST TIME: 01:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:21

SUBJECT'S LAST NAME: KLEPAC

DATE/TIME OF TEST: 01/01/2017 02:47:34

Diagnostics Check		OK	02:41
Air Blank		0.000	02:41
Control Test		0.081	02:42
Air Blank		0.000	02:42
Subject Sample	# 1	0.000	02:43
Breath Volume		1.382	02:43
Air Blank		0.000	02:43
Air Blank		0.000	02:45
Subject Sample	# 2	0.000	02:46
Breath Volume		2.074	02:46
Air Blank		0.000	02:46
Control Test		0.082	02:46
Air Blank		0.000	02:47
Diagnostics Check		OK	02:47

Data Download Date/Time: 01/12/2017 23:36

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEXANDER,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ELLIS,ALEXANDER
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG622905
DRY GAS EXPIRY DATE: 08/16/2018
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: EVANS

DATE/TIME OF TEST: 01/01/2017 03:57:41

Diagnostics Check		OK	03:51
Air Blank		0.000	03:52
Control Test		0.081	03:52
Air Blank		0.000	03:52
Subject Sample	# 1	0.205	03:53
Breath Volume		2.699	03:53
Air Blank		0.000	03:53
Air Blank		0.000	03:55
Subject Sample	# 2	0.189	03:56
Breath Volume		1.628	03:56
Air Blank		0.000	03:56
Control Test		0.081	03:57
Air Blank		0.000	03:57
Diagnostics Check		OK	03:57

Data Download Date/Time: 01/12/2017 23:36

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KASELAK,STEPHEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG622905
DRY GAS EXPIRY DATE: 08/16/2018
VIOLATION CODE: DUI
ARREST TIME: 04:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:38

SUBJECT'S LAST NAME: EMERY

DATE/TIME OF TEST: 01/01/2017 05:05:40

Diagnostics Check		OK	04:59
Air Blank		0.000	04:59
Control Test		0.081	04:59
Air Blank		0.000	05:00
Subject Sample	# 1	0.104	05:01
Breath Volume		3.070	05:01
Air Blank		0.000	05:01
Air Blank		0.000	05:03
Subject Sample	# 2	0.095	05:04
Breath Volume		2.437	05:04
Air Blank		0.000	05:04
Control Test		0.080	05:05
Air Blank		0.000	05:05
Diagnostics Check		OK	05:05

Data Download Date/Time: 01/12/2017 23:36

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ASHCOM,TIFFANY,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ASHCOM,TIFFANY
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG622905
DRY GAS EXPIRY DATE: 08/16/2018
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:13

SUBJECT'S LAST NAME: HAGUE

DATE/TIME OF TEST: 01/03/2017 01:35:57

Diagnostics Check		OK	01:33	
Air Blank		0.000	01:33	
Control Test		0.081	01:33	
Air Blank		0.000	01:34	
Subject Sample	# 1		01:34	Subject Test Refused
Breath Volume		0.000	01:34	
Air Blank		0.000	01:35	
Control Test		0.081	01:35	
Air Blank		0.000	01:35	
Diagnostics Check		OK	01:35	

Data Download Date/Time: 01/12/2017 23:36

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOWATT,NATHAN,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MOWATT,NATHAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG622905
DRY GAS EXPIRY DATE: 08/16/2018
VIOLATION CODE: DUI
ARREST TIME: 03:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:47

SUBJECT'S LAST NAME: CRAWFORD

DATE/TIME OF TEST: 01/03/2017 04:14:29

Diagnostics Check		OK	04:07
Air Blank		0.000	04:08
Control Test		0.081	04:08
Air Blank		0.000	04:09
Subject Sample	# 1	0.233	04:09
Breath Volume		2.394	04:09
Air Blank		0.000	04:10
Air Blank		0.000	04:12
Subject Sample	# 2	0.242	04:12
Breath Volume		3.527	04:12
Air Blank		0.000	04:13
Control Test		0.081	04:13
Air Blank		0.000	04:14
Diagnostics Check		OK	04:14

Data Download Date/Time: 01/12/2017 23:36

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOOD,COREY,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HOOD,COREY
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG622905
DRY GAS EXPIRY DATE: 08/16/2018
VIOLATION CODE: DUI
ARREST TIME: 15:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:25

SUBJECT'S LAST NAME: COOPER

DATE/TIME OF TEST: 01/03/2017 18:29:31

Diagnostics Check		OK	18:26	
Air Blank		0.000	18:27	
Control Test		0.082	18:27	
Air Blank		0.000	18:27	
Subject Sample	# 1		18:28	Subject Test Refused
Breath Volume		0.000	18:28	
Air Blank		0.000	18:28	
Control Test		0.082	18:28	
Air Blank		0.000	18:29	
Diagnostics Check		OK	18:29	

Data Download Date/Time: 01/12/2017 23:36

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEXANDER,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ELLIS,ALEXANDER
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG622905
DRY GAS EXPIRY DATE: 08/16/2018
VIOLATION CODE: DUI
ARREST TIME: 23:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: CALVERT

DATE/TIME OF TEST: 01/05/2017 00:44:23

Diagnostics Check		OK	00:37
Air Blank		0.000	00:38
Control Test		0.082	00:38
Air Blank		0.000	00:38
Subject Sample	# 1	0.173	00:39
Breath Volume		1.960	00:39
Air Blank		0.000	00:40
Air Blank		0.000	00:42
Subject Sample	# 2	0.184	00:42
Breath Volume		2.042	00:42
Air Blank		0.000	00:43
Control Test		0.081	00:43
Air Blank		0.000	00:44
Diagnostics Check		OK	00:44

Data Download Date/Time: 01/12/2017 23:36

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEXANDER,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ELLIS,ALEXANDER
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG622905
DRY GAS EXPIRY DATE: 08/16/2018
VIOLATION CODE: DUI
ARREST TIME: 02:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:09

SUBJECT'S LAST NAME: VITELLO

DATE/TIME OF TEST: 01/05/2017 03:40:58

Diagnostics Check		OK	03:34
Air Blank		0.000	03:35
Control Test		0.081	03:35
Air Blank		0.000	03:36
Subject Sample	# 1	0.151	03:36
Breath Volume		2.656	03:36
Air Blank		0.000	03:37
Air Blank		0.000	03:39
Subject Sample	# 2	0.148	03:39
Breath Volume		2.511	03:39
Air Blank		0.000	03:40
Control Test		0.082	03:40
Air Blank		0.000	03:40
Diagnostics Check		OK	03:40

Data Download Date/Time: 01/12/2017 23:36

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOBSON,JON,B
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: DOBSON,JON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG622905
DRY GAS EXPIRY DATE: 08/16/2018
VIOLATION CODE: DUI
ARREST TIME: 14:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:45

SUBJECT'S LAST NAME: ALLEN

DATE/TIME OF TEST: 01/05/2017 15:11:39

Diagnostics Check		OK	15:05
Air Blank		0.000	15:05
Control Test		0.082	15:06
Air Blank		0.000	15:06
Subject Sample	# 1	0.210	15:07
Breath Volume		2.308	15:07
Air Blank		0.000	15:07
Air Blank		0.000	15:09
Subject Sample	# 2	0.219	15:10
Breath Volume		1.753	15:10
Air Blank		0.000	15:10
Control Test		0.081	15:11
Air Blank		0.000	15:11
Diagnostics Check		OK	15:11

Data Download Date/Time: 01/12/2017 23:36

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEXANDER,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ELLIS,ALEXANDER
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG622905
DRY GAS EXPIRY DATE: 08/16/2018
VIOLATION CODE: DUI
ARREST TIME: 22:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:59

SUBJECT'S LAST NAME: ROLAND

DATE/TIME OF TEST: 01/05/2017 23:29:23

Diagnostics Check		OK	23:22
Air Blank		0.000	23:22
Control Test		0.082	23:23
Air Blank		0.000	23:23
Subject Sample	# 1	0.164	23:24
Breath Volume		1.394	23:24
Air Blank		0.000	23:24
Air Blank		0.000	23:26
Subject Sample	# 2	0.154	23:27
Breath Volume		1.234	23:27
Air Blank		0.000	23:28
Control Test		0.082	23:28
Air Blank		0.000	23:29
Diagnostics Check		OK	23:29

Data Download Date/Time: 01/12/2017 23:36

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALATI,VINCENZO,F
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GALATI,VINCENZO
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG622905
DRY GAS EXPIRY DATE: 08/16/2018
VIOLATION CODE: DUI
ARREST TIME: 23:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:18

SUBJECT'S LAST NAME: O BRIEN

DATE/TIME OF TEST: 01/06/2017 00:50:51

Diagnostics Check		OK	00:44
Air Blank		0.000	00:45
Control Test		0.082	00:45
Air Blank		0.000	00:45
Subject Sample	# 1	0.000	00:46
Breath Volume		1.847	00:46
Air Blank		0.000	00:46
Air Blank		0.000	00:48
Subject Sample	# 2	0.000	00:49
Breath Volume		1.238	00:49
Air Blank		0.000	00:49
Control Test		0.082	00:50
Air Blank		0.000	00:50
Diagnostics Check		OK	00:50

Data Download Date/Time: 01/12/2017 23:36

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEXANDER,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ELLIS,ALEXANDER
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG622905
DRY GAS EXPIRY DATE: 08/16/2018
VIOLATION CODE: DUI
ARREST TIME: 01:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:23

SUBJECT'S LAST NAME: FLETCHER

DATE/TIME OF TEST: 01/07/2017 02:46:49

Diagnostics Check		OK	02:43	
Air Blank		0.000	02:44	
Control Test		0.082	02:44	
Air Blank		0.000	02:44	
Subject Sample	# 1		02:45	RFI Detect
Breath Volume		0.000	02:45	
Air Blank		0.000	02:45	
Control Test		0.081	02:46	
Air Blank		0.000	02:46	
Diagnostics Check		OK	02:46	

Data Download Date/Time: 01/12/2017 23:37

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEXANDER,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ELLIS,ALEXANDER
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG622905
DRY GAS EXPIRY DATE: 08/16/2018
VIOLATION CODE: DUI
ARREST TIME: 01:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:23

SUBJECT'S LAST NAME: FLETCHER

DATE/TIME OF TEST: 01/07/2017 02:51:39

Diagnostics Check		OK	02:48	
Air Blank		0.000	02:49	
Control Test		0.082	02:49	
Air Blank		0.000	02:50	
Subject Sample	# 1		02:50	Subject Test Refused
Breath Volume		0.000	02:50	
Air Blank		0.000	02:50	
Control Test		0.081	02:51	
Air Blank		0.000	02:51	
Diagnostics Check		OK	02:51	

Data Download Date/Time: 01/12/2017 23:37

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALATI,VINCENZO,F
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GALATI,VINCENZO
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG622905
DRY GAS EXPIRY DATE: 08/16/2018
VIOLATION CODE: DUI
ARREST TIME: 02:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:12

SUBJECT'S LAST NAME: BAGNALL

DATE/TIME OF TEST: 01/07/2017 03:48:16

Diagnostics Check		OK	03:39
Air Blank		0.000	03:40
Control Test		0.082	03:40
Air Blank		0.000	03:41
Subject Sample	# 1	0.044	03:43
Breath Volume		1.589	03:43
Air Blank		0.000	03:44
Air Blank		0.000	03:46
Subject Sample	# 2	0.041	03:46
Breath Volume		1.562	03:46
Air Blank		0.000	03:47
Control Test		0.081	03:47
Air Blank		0.000	03:48
Diagnostics Check		OK	03:48

Data Download Date/Time: 01/12/2017 23:37

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOWATT,NATHAN,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MOWATT,NATHAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG622905
DRY GAS EXPIRY DATE: 08/16/2018
VIOLATION CODE: DUI
ARREST TIME: 05:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:36

SUBJECT'S LAST NAME: CHABRIKOV

DATE/TIME OF TEST: 01/07/2017 06:04:03

Diagnostics Check		OK	05:56
Air Blank		0.000	05:57
Control Test		0.082	05:57
Air Blank		0.000	05:57
Subject Sample	# 1	0.172	05:58
Breath Volume		1.132	05:58
Air Blank		0.000	05:58
Air Blank		0.000	06:00
Subject Sample	# 2	0.162	06:02
Breath Volume		1.804	06:02
Air Blank		0.000	06:03
Control Test		0.081	06:03
Air Blank		0.000	06:03
Diagnostics Check		OK	06:04

Data Download Date/Time: 01/12/2017 23:37

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAIR,RONALD,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAIR,RONALD
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG622905
DRY GAS EXPIRY DATE: 08/16/2018
VIOLATION CODE: DUI
ARREST TIME: 23:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:46

SUBJECT'S LAST NAME: REEVES

DATE/TIME OF TEST: 01/08/2017 00:11:01

Diagnostics Check		OK	00:07	
Air Blank		0.000	00:07	
Control Test		0.082	00:07	
Air Blank		0.000	00:08	
Subject Sample	# 1		00:09	Subject Test Refused
Breath Volume		0.000	00:09	
Air Blank		0.000	00:10	
Control Test		0.082	00:10	
Air Blank		0.000	00:10	
Diagnostics Check		OK	00:11	

Data Download Date/Time: 01/12/2017 23:37

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOWATT,NATHAN,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MOWATT,NATHAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG622905
DRY GAS EXPIRY DATE: 08/16/2018
VIOLATION CODE: DUI
ARREST TIME: 01:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:06

SUBJECT'S LAST NAME: DEMAREST

DATE/TIME OF TEST: 01/08/2017 02:32:55

Diagnostics Check		OK	02:26
Air Blank		0.000	02:26
Control Test		0.081	02:27
Air Blank		0.000	02:27
Subject Sample	# 1	0.245	02:28
Breath Volume		2.199	02:28
Air Blank		0.000	02:28
Air Blank		0.000	02:30
Subject Sample	# 2	0.240	02:31
Breath Volume		1.875	02:31
Air Blank		0.000	02:32
Control Test		0.082	02:32
Air Blank		0.000	02:32
Diagnostics Check		OK	02:32

Data Download Date/Time: 01/12/2017 23:37

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALATI,VINCENZO,F
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GALATI,VINCENZO
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG622905
DRY GAS EXPIRY DATE: 08/16/2018
VIOLATION CODE: DUI
ARREST TIME: 03:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:58

SUBJECT'S LAST NAME: NORRIS

DATE/TIME OF TEST: 01/08/2017 04:36:24

Diagnostics Check		OK	04:33	
Air Blank		0.000	04:33	
Control Test		0.083	04:34	
Air Blank		0.000	04:34	
Subject Sample	# 1		04:35	Subject Test Refused
Breath Volume		0.000	04:35	
Air Blank		0.000	04:35	
Control Test		0.083	04:35	
Air Blank		0.000	04:36	
Diagnostics Check		OK	04:36	

Data Download Date/Time: 01/12/2017 23:37

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOWATT,NATHAN,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MOWATT,NATHAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG622905
DRY GAS EXPIRY DATE: 08/16/2018
VIOLATION CODE: DUI
ARREST TIME: 02:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:21

SUBJECT'S LAST NAME: ROMAN

DATE/TIME OF TEST: 01/09/2017 03:49:36

Diagnostics Check		OK	03:41
Air Blank		0.000	03:42
Control Test		0.083	03:42
Air Blank		0.000	03:43
Subject Sample	# 1	0.187	03:44
Breath Volume		1.261	03:44
Air Blank		0.000	03:44
Air Blank		0.000	03:46
Subject Sample	# 2	0.188	03:48
Breath Volume		1.101	03:48
Air Blank		0.000	03:48
Control Test		0.079	03:49
Air Blank		0.000	03:49
Diagnostics Check		OK	03:49

Data Download Date/Time: 01/12/2017 23:37

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOWATT,NATHAN,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: PROPHETE,ELIE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG622905
DRY GAS EXPIRY DATE: 08/16/2018
VIOLATION CODE: DUI
ARREST TIME: 03:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:06

SUBJECT'S LAST NAME: MAISONET

DATE/TIME OF TEST: 01/09/2017 04:33:14

Diagnostics Check		OK	04:26
Air Blank		0.000	04:27
Control Test		0.082	04:27
Air Blank		0.000	04:28
Subject Sample	# 1	0.187	04:28
Breath Volume		1.281	04:28
Air Blank		0.000	04:29
Air Blank		0.000	04:31
Subject Sample	# 2	0.188	04:31
Breath Volume		1.632	04:31
Air Blank		0.000	04:32
Control Test		0.081	04:32
Air Blank		0.000	04:33
Diagnostics Check		OK	04:33

Data Download Date/Time: 01/12/2017 23:37

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOWATT,NATHAN,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MOWATT,NATHAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG622905
DRY GAS EXPIRY DATE: 08/16/2018
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: DRECHSEL

DATE/TIME OF TEST: 01/10/2017 01:38:07

Diagnostics Check		OK	01:30
Air Blank		0.000	01:31
Control Test		0.082	01:31
Air Blank		0.000	01:31
Subject Sample	# 1	0.258	01:33
Breath Volume		3.457	01:33
Air Blank		0.000	01:33
Air Blank		0.000	01:35
Subject Sample	# 2	0.257	01:36
Breath Volume		3.328	01:36
Air Blank		0.000	01:37
Control Test		0.078	01:37
Air Blank		0.000	01:38
Diagnostics Check		OK	01:38

Data Download Date/Time: 01/12/2017 23:37

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DORR,RANDALL,S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: DORR,RANDALL
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG622905
DRY GAS EXPIRY DATE: 08/16/2018
VIOLATION CODE: DUI
ARREST TIME: 03:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:25

SUBJECT'S LAST NAME: MIDDLETON

DATE/TIME OF TEST: 01/10/2017 04:57:15

Diagnostics Check		OK	04:45	
Air Blank		0.000	04:45	
Control Test		0.082	04:45	
Air Blank		0.000	04:46	
Subject Sample	# 1	0.192	04:47	
Breath Volume		1.472	04:47	
Air Blank		0.000	04:48	
Air Blank		0.000	04:49	
Subject Sample	# 2		04:52	No Sample Provided
Breath Volume		0.000	04:52	
Air Blank		0.000	04:53	
Air Blank		0.000	04:55	
Subject Sample	# 3		04:55	Subject Test Refused
Breath Volume		0.000	04:55	
Air Blank		0.000	04:56	
Control Test		0.082	04:56	
Air Blank		0.000	04:57	
Diagnostics Check		OK	04:57	

Data Download Date/Time: 01/12/2017 23:37

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOOD,COREY,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HOOD,COREY
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG622905
DRY GAS EXPIRY DATE: 08/16/2018
VIOLATION CODE: DUI
ARREST TIME: 20:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:15

SUBJECT'S LAST NAME: LENNEN

DATE/TIME OF TEST: 01/10/2017 21:43:33

Diagnostics Check		OK	21:36
Air Blank		0.000	21:36
Control Test		0.082	21:36
Air Blank		0.000	21:37
Subject Sample	# 1	0.198	21:38
Breath Volume		1.761	21:38
Air Blank		0.000	21:38
Air Blank		0.000	21:40
Subject Sample	# 2	0.187	21:41
Breath Volume		1.597	21:41
Air Blank		0.000	21:42
Control Test		0.082	21:42
Air Blank		0.000	21:43
Diagnostics Check		OK	21:43

Data Download Date/Time: 01/12/2017 23:37

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEXANDER,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ELLIS,ALEXANDER
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG622905
DRY GAS EXPIRY DATE: 08/16/2018
VIOLATION CODE: DUI
ARREST TIME: 23:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: POLSON

DATE/TIME OF TEST: 01/12/2017 00:49:36

Diagnostics Check		OK	00:45	
Air Blank		0.000	00:45	
Control Test		0.082	00:46	
Air Blank		0.000	00:46	
Subject Sample	# 1		00:48	Subject Test Refused
Breath Volume		0.000	00:48	
Air Blank		0.000	00:48	
Control Test		0.081	00:49	
Air Blank		0.000	00:49	
Diagnostics Check		OK	00:49	

Data Download Date/Time: 01/12/2017 23:37

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEXANDER,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ELLIS,ALEX
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG622905
DRY GAS EXPIRY DATE: 08/16/2018
VIOLATION CODE: DUI
ARREST TIME: 03:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:32

SUBJECT'S LAST NAME: CLARK

DATE/TIME OF TEST: 01/12/2017 04:00:40

Diagnostics Check		OK	03:54
Air Blank		0.000	03:54
Control Test		0.081	03:54
Air Blank		0.000	03:55
Subject Sample	# 1	0.150	03:56
Breath Volume		1.546	03:56
Air Blank		0.000	03:56
Air Blank		0.000	03:58
Subject Sample	# 2	0.159	03:59
Breath Volume		1.714	03:59
Air Blank		0.000	03:59
Control Test		0.080	04:00
Air Blank		0.000	04:00
Diagnostics Check		OK	04:00

Data Download Date/Time: 01/12/2017 23:37

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KASELAK,STEPHEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG622905
DRY GAS EXPIRY DATE: 08/16/2018
VIOLATION CODE: DUI
ARREST TIME: 20:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:52

SUBJECT'S LAST NAME: BROWNLEE

DATE/TIME OF TEST: 01/12/2017 21:22:01

Diagnostics Check		OK	21:12	
Air Blank		0.000	21:13	
Control Test		0.081	21:13	
Air Blank		0.000	21:13	
Subject Sample	# 1		21:17	No Sample Provided
Breath Volume		0.000	21:17	
Air Blank		0.000	21:17	
Air Blank		0.000	21:19	
Subject Sample	# 2		21:20	Subject Test Refused
Breath Volume		0.000	21:20	
Air Blank		0.000	21:21	
Control Test		0.082	21:21	
Air Blank		0.000	21:21	
Diagnostics Check		OK	21:22	

Data Download Date/Time: 01/12/2017 23:37

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALATI,VINCENZO,F
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GALATI,VINCENZO
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG622905
DRY GAS EXPIRY DATE: 08/16/2018
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:16

SUBJECT'S LAST NAME: CULLEN

DATE/TIME OF TEST: 01/13/2017 03:39:44

Diagnostics Check		OK	03:36	
Air Blank		0.000	03:37	
Control Test		0.081	03:37	
Air Blank		0.000	03:38	
Subject Sample	# 1		03:38	Subject Test Refused
Breath Volume		0.000	03:38	
Air Blank		0.000	03:38	
Control Test		0.081	03:39	
Air Blank		0.000	03:39	
Diagnostics Check		OK	03:39	

Data Download Date/Time: 02/03/2017 00:09

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEXANDER,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ELLIS,ALEXANDER
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG622905
DRY GAS EXPIRY DATE: 08/16/2018
VIOLATION CODE: DUI
ARREST TIME: 20:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: BACH

DATE/TIME OF TEST: 01/13/2017 21:34:13

Diagnostics Check		OK	21:30	
Air Blank		0.000	21:31	
Control Test		0.081	21:31	
Air Blank		0.000	21:31	
Subject Sample	# 1		21:32	Subject Test Refused
Breath Volume		0.000	21:32	
Air Blank		0.000	21:33	
Control Test		0.081	21:33	
Air Blank		0.000	21:34	
Diagnostics Check		OK	21:34	

Data Download Date/Time: 02/03/2017 00:09

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALATI,VINCENZO,F
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GALATI,VINCENZO
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG622905
DRY GAS EXPIRY DATE: 08/16/2018
VIOLATION CODE: DUI
ARREST TIME: 02:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:58

SUBJECT'S LAST NAME: KUZEL

DATE/TIME OF TEST: 01/14/2017 03:28:44

Diagnostics Check		OK	03:25	
Air Blank		0.000	03:25	
Control Test		0.081	03:26	
Air Blank		0.000	03:26	
Subject Sample	# 1		03:27	Subject Test Refused
Breath Volume		0.000	03:27	
Air Blank		0.000	03:27	
Control Test		0.081	03:28	
Air Blank		0.000	03:28	
Diagnostics Check		OK	03:28	

Data Download Date/Time: 02/03/2017 00:09

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEXANDER,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ELLIS,A
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG622905
DRY GAS EXPIRY DATE: 08/16/2018
VIOLATION CODE: DUI
ARREST TIME: 03:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: MCLAUGHLIN

DATE/TIME OF TEST: 01/14/2017 04:54:08

Diagnostics Check		OK	04:43	
Air Blank		0.000	04:44	
Control Test		0.081	04:44	
Air Blank		0.000	04:44	
Subject Sample	# 1		04:48	No Sample Provided
Breath Volume		0.000	04:48	
Air Blank		0.000	04:48	
Air Blank		0.000	04:50	
Subject Sample	# 2		04:52	Subject Test Refused
Breath Volume		0.000	04:52	
Air Blank		0.000	04:53	
Control Test		0.081	04:53	
Air Blank		0.000	04:54	
Diagnostics Check		OK	04:54	

Data Download Date/Time: 02/03/2017 00:09

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SWANNER,SARAH,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LOCKE,REBECCA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG622905
DRY GAS EXPIRY DATE: 08/16/2018
VIOLATION CODE: DUI
ARREST TIME: 17:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:11

SUBJECT'S LAST NAME: JACOBS

DATE/TIME OF TEST: 01/14/2017 18:39:45

Diagnostics Check		OK	18:31
Air Blank		0.000	18:31
Control Test		0.082	18:32
Air Blank		0.000	18:32
Subject Sample	# 1	0.364	18:35
Breath Volume		2.054	18:35
Air Blank		0.000	18:35
Air Blank		0.000	18:37
Subject Sample	# 2	0.353	18:38
Breath Volume		1.699	18:38
Air Blank		0.000	18:38
Control Test		0.078	18:39
Air Blank		0.000	18:39
Diagnostics Check		OK	18:39

Data Download Date/Time: 02/03/2017 00:09

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEXANDER,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ELLIS,ALEXANDER
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG622905
DRY GAS EXPIRY DATE: 08/16/2018
VIOLATION CODE: DUI
ARREST TIME: 23:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: GRZEGORZEWSKI

DATE/TIME OF TEST: 01/15/2017 00:00:01

Diagnostics Check		OK	23:52
Air Blank		0.000	23:52
Control Test		0.081	23:52
Air Blank		0.000	23:53
Subject Sample	# 1	0.109	23:55
Breath Volume		1.367	23:55
Air Blank		0.000	23:55
Air Blank		0.000	23:57
Subject Sample	# 2	0.104	23:58
Breath Volume		1.488	23:58
Air Blank		0.000	23:59
Control Test		0.081	23:59
Air Blank		0.000	23:59
Diagnostics Check		OK	00:00

Data Download Date/Time: 02/03/2017 00:09

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: CASEY,K
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG622905
DRY GAS EXPIRY DATE: 08/16/2018
VIOLATION CODE: DUI
ARREST TIME: 22:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: HAYNES

DATE/TIME OF TEST: 01/18/2017 23:33:11

Diagnostics Check		OK	23:20	
Air Blank		0.000	23:21	
Control Test		0.082	23:21	
Air Blank		0.000	23:22	
Subject Sample	# 1	0.120	23:25	Volume Not Met
Breath Volume		0.363	23:25	
Air Blank		0.000	23:26	
Air Blank		0.000	23:27	
Subject Sample	# 2	0.145	23:28	
Breath Volume		1.730	23:28	
Air Blank		0.000	23:29	
Air Blank		0.000	23:30	
Subject Sample	# 3	0.147	23:31	
Breath Volume		1.335	23:31	
Air Blank		0.000	23:32	
Control Test		0.078	23:32	
Air Blank		0.000	23:33	
Diagnostics Check		OK	23:33	

Data Download Date/Time: 02/03/2017 00:09

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEXANDER,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ELLIS,ALEXANDER
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG622905
DRY GAS EXPIRY DATE: 08/16/2018
VIOLATION CODE: DUI
ARREST TIME: 04:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:47

SUBJECT'S LAST NAME: SEXTON

DATE/TIME OF TEST: 01/19/2017 05:14:24

Diagnostics Check		OK	05:10	
Air Blank		0.000	05:10	
Control Test		0.082	05:11	
Air Blank		0.000	05:11	
Subject Sample	# 1		05:13	Subject Test Refused
Breath Volume		0.000	05:13	
Air Blank		0.000	05:13	
Control Test		0.081	05:13	
Air Blank		0.000	05:14	
Diagnostics Check		OK	05:14	

Data Download Date/Time: 02/03/2017 00:09

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: DONOHUE,S
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG622905
DRY GAS EXPIRY DATE: 08/16/2018
VIOLATION CODE: DUI
ARREST TIME: 19:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:45

SUBJECT'S LAST NAME: CHURCHILL

DATE/TIME OF TEST: 01/19/2017 21:09:51

Diagnostics Check		OK	21:05	
Air Blank		0.000	21:06	
Control Test		0.082	21:06	
Air Blank		0.000	21:06	
Subject Sample	# 1		21:08	Subject Test Refused
Breath Volume		0.000	21:08	
Air Blank		0.000	21:08	
Control Test		0.081	21:09	
Air Blank		0.000	21:09	
Diagnostics Check		OK	21:09	

Data Download Date/Time: 02/03/2017 00:09

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEXANDER,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ELLIS,ALEXANDER
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG622905
DRY GAS EXPIRY DATE: 08/16/2018
VIOLATION CODE: DUI
ARREST TIME: 01:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: BRAUN

DATE/TIME OF TEST: 01/20/2017 03:01:02

Diagnostics Check		OK	02:50
Air Blank		0.000	02:51
Control Test		0.082	02:51
Air Blank		0.000	02:52
Subject Sample	# 1	0.000	02:54
Breath Volume		1.453	02:54
Air Blank		0.000	02:55
Air Blank		0.000	02:57
Subject Sample	# 2	0.000	02:59
Breath Volume		1.242	02:59
Air Blank		0.000	03:00
Control Test		0.081	03:00
Air Blank		0.000	03:00
Diagnostics Check		OK	03:01

Data Download Date/Time: 02/03/2017 00:09

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEXANDER,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ELLIS,ALEXANDER
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG622905
DRY GAS EXPIRY DATE: 08/16/2018
VIOLATION CODE: DUI
ARREST TIME: 22:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: SCUDAMORE

DATE/TIME OF TEST: 01/20/2017 23:27:51

Diagnostics Check		OK	23:20	
Air Blank		0.000	23:21	
Control Test		0.083	23:21	
Air Blank		0.000	23:22	
Subject Sample	# 1	0.000	23:23	
Breath Volume		1.785	23:23	
Air Blank		0.000	23:24	
Air Blank		0.000	23:26	
Subject Sample	# 2		23:26	Subject Test Refused
Breath Volume		0.000	23:26	
Air Blank		0.000	23:26	
Control Test		0.082	23:27	
Air Blank		0.000	23:27	
Diagnostics Check		OK	23:27	

Data Download Date/Time: 02/03/2017 00:09

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEXANDER,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ELLIS,A
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG622905
DRY GAS EXPIRY DATE: 08/16/2018
VIOLATION CODE: DUI
ARREST TIME: 00:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:59

SUBJECT'S LAST NAME: MAGNUSON

DATE/TIME OF TEST: 01/21/2017 01:26:13

Diagnostics Check		OK	01:20
Air Blank		0.000	01:20
Control Test		0.081	01:20
Air Blank		0.000	01:21
Subject Sample	# 1	0.206	01:21
Breath Volume		1.835	01:21
Air Blank		0.000	01:22
Air Blank		0.000	01:24
Subject Sample	# 2	0.211	01:24
Breath Volume		1.761	01:24
Air Blank		0.000	01:25
Control Test		0.081	01:25
Air Blank		0.000	01:26
Diagnostics Check		OK	01:26

Data Download Date/Time: 02/03/2017 00:09

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WELSHANS,DEAN,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WELSHANS,DEAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG622905
DRY GAS EXPIRY DATE: 08/16/2018
VIOLATION CODE: DUI
ARREST TIME: 02:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:42

SUBJECT'S LAST NAME: PASSARELLA

DATE/TIME OF TEST: 01/21/2017 03:09:27

Diagnostics Check		OK	03:02
Air Blank		0.000	03:03
Control Test		0.082	03:03
Air Blank		0.000	03:04
Subject Sample	# 1	0.175	03:04
Breath Volume		4.105	03:04
Air Blank		0.000	03:05
Air Blank		0.000	03:07
Subject Sample	# 2	0.181	03:07
Breath Volume		4.398	03:07
Air Blank		0.000	03:08
Control Test		0.081	03:08
Air Blank		0.000	03:09
Diagnostics Check		OK	03:09

Data Download Date/Time: 02/03/2017 00:09

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BREGON,JAMES,S
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: BREGON,JAMES
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG622905
DRY GAS EXPIRY DATE: 08/16/2018
VIOLATION CODE: DUI
ARREST TIME: 02:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:27

SUBJECT'S LAST NAME: CINELLI

DATE/TIME OF TEST: 01/21/2017 04:50:15

Diagnostics Check		OK	04:47	
Air Blank		0.000	04:47	
Control Test		0.081	04:47	
Air Blank		0.000	04:48	
Subject Sample	# 1		04:48	Subject Test Refused
Breath Volume		0.000	04:48	
Air Blank		0.000	04:49	
Control Test		0.081	04:49	
Air Blank		0.000	04:50	
Diagnostics Check		OK	04:50	

Data Download Date/Time: 02/03/2017 00:09

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MACKAIL,B
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG622905
DRY GAS EXPIRY DATE: 08/16/2018
VIOLATION CODE: DUI
ARREST TIME: 23:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:24

SUBJECT'S LAST NAME: SHIRAH

DATE/TIME OF TEST: 01/22/2017 00:47:44

Diagnostics Check		OK	00:44	
Air Blank		0.000	00:45	
Control Test		0.081	00:45	
Air Blank		0.000	00:46	
Subject Sample	# 1		00:46	Subject Test Refused
Breath Volume		0.000	00:46	
Air Blank		0.000	00:46	
Control Test		0.080	00:47	
Air Blank		0.000	00:47	
Diagnostics Check		OK	00:47	

Data Download Date/Time: 02/03/2017 00:09

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MIDDLETON,KRISTEN,M
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: MIDDLETON,KRISTEN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG622905
DRY GAS EXPIRY DATE: 08/16/2018
VIOLATION CODE: DUI
ARREST TIME: 15:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:44

SUBJECT'S LAST NAME: ESKELSON-KEENAN

DATE/TIME OF TEST: 01/24/2017 17:07:17

Diagnostics Check		OK	17:04	
Air Blank		0.000	17:04	
Control Test		0.082	17:05	
Air Blank		0.000	17:05	
Subject Sample	# 1		17:05	Subject Test Refused
Breath Volume		0.000	17:05	
Air Blank		0.000	17:06	
Control Test		0.082	17:06	
Air Blank		0.000	17:07	
Diagnostics Check		OK	17:07	

Data Download Date/Time: 02/03/2017 00:09

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DORR,RANDALL,S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: DORR,RANDALL
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG622905
DRY GAS EXPIRY DATE: 08/16/2018
VIOLATION CODE: DUI
ARREST TIME: 02:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:04

SUBJECT'S LAST NAME: MORRIS

DATE/TIME OF TEST: 01/26/2017 03:39:52

Diagnostics Check		OK	03:35	
Air Blank		0.000	03:36	
Control Test		0.082	03:36	
Air Blank		0.000	03:37	
Subject Sample	# 1		03:38	Subject Test Refused
Breath Volume		0.000	03:38	
Air Blank		0.000	03:38	
Control Test		0.082	03:39	
Air Blank		0.000	03:39	
Diagnostics Check		OK	03:39	

Data Download Date/Time: 02/03/2017 00:09

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LANEY,DAMON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG622905
DRY GAS EXPIRY DATE: 08/16/2018
VIOLATION CODE: DUI
ARREST TIME: 03:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 01/26/2017 04:15:33

Diagnostics Check		OK	04:09
Air Blank		0.000	04:09
Control Test		0.081	04:10
Air Blank		0.000	04:10
Subject Sample	# 1	0.167	04:11
Breath Volume		1.304	04:11
Air Blank		0.000	04:11
Air Blank		0.000	04:13
Subject Sample	# 2	0.176	04:13
Breath Volume		1.304	04:13
Air Blank		0.000	04:14
Control Test		0.081	04:14
Air Blank		0.000	04:15
Diagnostics Check		OK	04:15

Data Download Date/Time: 02/03/2017 00:09

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALATI,VINCENZO,F
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GALATI,VINCENZO
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG622905
DRY GAS EXPIRY DATE: 08/16/2018
VIOLATION CODE: DUI
ARREST TIME: 04:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:14

SUBJECT'S LAST NAME: MCFADDEN

DATE/TIME OF TEST: 01/26/2017 05:40:51

Diagnostics Check		OK	05:34
Air Blank		0.000	05:34
Control Test		0.082	05:35
Air Blank		0.000	05:35
Subject Sample	# 1	0.112	05:36
Breath Volume		1.660	05:36
Air Blank		0.000	05:36
Air Blank		0.000	05:38
Subject Sample	# 2	0.118	05:39
Breath Volume		3.312	05:39
Air Blank		0.000	05:39
Control Test		0.082	05:40
Air Blank		0.000	05:40
Diagnostics Check		OK	05:40

Data Download Date/Time: 02/03/2017 00:09

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAIR,RONALD,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAIR,RONALD
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG622905
DRY GAS EXPIRY DATE: 08/16/2018
VIOLATION CODE: DUI
ARREST TIME: 19:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:18

SUBJECT'S LAST NAME: ORRICO

DATE/TIME OF TEST: 01/26/2017 20:44:21

Diagnostics Check		OK	20:38
Air Blank		0.000	20:38
Control Test		0.081	20:39
Air Blank		0.000	20:39
Subject Sample	# 1	0.000	20:40
Breath Volume		2.628	20:40
Air Blank		0.000	20:40
Air Blank		0.000	20:42
Subject Sample	# 2	0.000	20:42
Breath Volume		2.132	20:42
Air Blank		0.000	20:43
Control Test		0.080	20:43
Air Blank		0.000	20:44
Diagnostics Check		OK	20:44

Data Download Date/Time: 02/03/2017 00:09

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUARTE,CHRISTOPHER,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: DUARTE,CHRISTOPHER
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/12/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: AG622905
DRY GAS EXPIRY DATE: 08/16/2018
VIOLATION CODE: DUI
ARREST TIME: 21:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:25

SUBJECT'S LAST NAME: KOFOED

DATE/TIME OF TEST: 01/26/2017 22:50:10

Diagnostics Check		OK	22:46	
Air Blank		0.000	22:47	
Control Test		0.080	22:47	
Air Blank		0.000	22:48	
Subject Sample	# 1		22:48	Subject Test Refused
Breath Volume		0.000	22:48	
Air Blank		0.000	22:49	
Control Test		0.080	22:49	
Air Blank		0.000	22:50	
Diagnostics Check		OK	22:50	

Data Download Date/Time: 02/03/2017 00:09

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEXANDER,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ELLIS,ALEXANDER
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG622905
DRY GAS EXPIRY DATE: 08/16/2018
VIOLATION CODE: DUI
ARREST TIME: 01:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: VEREB

DATE/TIME OF TEST: 01/27/2017 02:54:12

Diagnostics Check		OK	02:51	
Air Blank		0.000	02:51	
Control Test		0.081	02:52	
Air Blank		0.000	02:52	
Subject Sample	# 1		02:52	RFI Detect
Breath Volume		0.000	02:52	
Air Blank		0.000	02:53	
Control Test		0.081	02:53	
Air Blank		0.000	02:54	
Diagnostics Check		OK	02:54	

Data Download Date/Time: 02/03/2017 00:09

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEXANDER,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ELLIS,ALEXANDER
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG622905
DRY GAS EXPIRY DATE: 08/16/2018
VIOLATION CODE: DUI
ARREST TIME: 01:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:14

SUBJECT'S LAST NAME: VEREB

DATE/TIME OF TEST: 01/27/2017 02:59:09

Diagnostics Check		OK	02:56	
Air Blank		0.000	02:56	
Control Test		0.081	02:57	
Air Blank		0.000	02:57	
Subject Sample	# 1		02:57	Subject Test Refused
Breath Volume		0.000	02:57	
Air Blank		0.000	02:58	
Control Test		0.080	02:58	
Air Blank		0.000	02:59	
Diagnostics Check		OK	02:59	

Data Download Date/Time: 02/03/2017 00:09

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LANEY,DAMON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG622905
DRY GAS EXPIRY DATE: 08/16/2018
VIOLATION CODE: DUI
ARREST TIME: 03:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: ZVOSECHZ

DATE/TIME OF TEST: 01/27/2017 04:48:14

Diagnostics Check		OK	04:40
Air Blank		0.000	04:41
Control Test		0.081	04:41
Air Blank		0.000	04:42
Subject Sample	# 1	0.128	04:42
Breath Volume		1.101	04:42
Air Blank		0.000	04:43
Air Blank		0.000	04:45
Subject Sample	# 2	0.126	04:46
Breath Volume		1.578	04:46
Air Blank		0.000	04:47
Control Test		0.081	04:47
Air Blank		0.000	04:48
Diagnostics Check		OK	04:48

Data Download Date/Time: 02/03/2017 00:09

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALATI,VINCENZO,F
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GALATI,VINCENZO
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG622905
DRY GAS EXPIRY DATE: 08/16/2018
VIOLATION CODE: DUI
ARREST TIME: 02:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:36

SUBJECT'S LAST NAME: CARVAJAL SANTANA

DATE/TIME OF TEST: 01/28/2017 03:08:47

Diagnostics Check		OK	03:02
Air Blank		0.000	03:02
Control Test		0.081	03:02
Air Blank		0.000	03:03
Subject Sample	# 1	0.164	03:03
Breath Volume		2.062	03:03
Air Blank		0.000	03:04
Air Blank		0.000	03:06
Subject Sample	# 2	0.169	03:07
Breath Volume		1.796	03:07
Air Blank		0.000	03:07
Control Test		0.082	03:08
Air Blank		0.000	03:08
Diagnostics Check		OK	03:08

Data Download Date/Time: 02/03/2017 00:09

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEXANDER,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ELLIS,ALEXANDER
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG622905
DRY GAS EXPIRY DATE: 08/16/2018
VIOLATION CODE: DUI
ARREST TIME: 03:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:49

SUBJECT'S LAST NAME: ROSS

DATE/TIME OF TEST: 01/28/2017 04:13:12

Diagnostics Check		OK	04:10	
Air Blank		0.000	04:10	
Control Test		0.082	04:11	
Air Blank		0.000	04:11	
Subject Sample	# 1		04:11	Subject Test Refused
Breath Volume		0.000	04:11	
Air Blank		0.000	04:12	
Control Test		0.082	04:12	
Air Blank		0.000	04:13	
Diagnostics Check		OK	04:13	

Data Download Date/Time: 02/03/2017 00:09

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALATI,VINCENZO,F
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GALATI,VINCENZO
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG622905
DRY GAS EXPIRY DATE: 08/16/2018
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:53

SUBJECT'S LAST NAME: LAHUTI

DATE/TIME OF TEST: 01/29/2017 01:19:47

Diagnostics Check		OK	01:13
Air Blank		0.000	01:14
Control Test		0.081	01:14
Air Blank		0.000	01:14
Subject Sample	# 1	0.068	01:15
Breath Volume		2.304	01:15
Air Blank		0.000	01:16
Air Blank		0.000	01:17
Subject Sample	# 2	0.060	01:18
Breath Volume		1.820	01:18
Air Blank		0.000	01:18
Control Test		0.081	01:19
Air Blank		0.000	01:19
Diagnostics Check		OK	01:19

Data Download Date/Time: 02/03/2017 00:09

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUARTE,CHRISTOPHER,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: DUARTE,CHRISTOPHER
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG622905
DRY GAS EXPIRY DATE: 08/16/2018
VIOLATION CODE: DUI
ARREST TIME: 05:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:53

SUBJECT'S LAST NAME: STILLWELL

DATE/TIME OF TEST: 01/29/2017 07:16:42

Diagnostics Check		OK	07:13	
Air Blank		0.000	07:14	
Control Test		0.082	07:14	
Air Blank		0.000	07:15	
Subject Sample	# 1		07:15	Subject Test Refused
Breath Volume		0.000	07:15	
Air Blank		0.000	07:15	
Control Test		0.082	07:16	
Air Blank		0.000	07:16	
Diagnostics Check		OK	07:16	

Data Download Date/Time: 02/03/2017 00:09

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ASHCOM,TIFFANY,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LEFFINGWELL,BRADLEY
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG622905
DRY GAS EXPIRY DATE: 08/16/2018
VIOLATION CODE: DUI
ARREST TIME: 21:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:18

SUBJECT'S LAST NAME: HART

DATE/TIME OF TEST: 01/29/2017 22:45:02

Diagnostics Check		OK	22:38
Air Blank		0.000	22:39
Control Test		0.082	22:39
Air Blank		0.000	22:40
Subject Sample	# 1	0.055	22:40
Breath Volume		2.781	22:40
Air Blank		0.000	22:41
Air Blank		0.000	22:43
Subject Sample	# 2	0.057	22:43
Breath Volume		2.945	22:43
Air Blank		0.000	22:44
Control Test		0.082	22:44
Air Blank		0.000	22:44
Diagnostics Check		OK	22:45

Data Download Date/Time: 02/03/2017 00:09

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ASHCOM,TIFFANY,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TEMPLE,ELIZABETH
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG622905
DRY GAS EXPIRY DATE: 08/16/2018
VIOLATION CODE: DUI
ARREST TIME: 22:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:52

SUBJECT'S LAST NAME: ALBERTY

DATE/TIME OF TEST: 01/29/2017 23:15:10

Diagnostics Check		OK	23:12	
Air Blank		0.000	23:12	
Control Test		0.082	23:13	
Air Blank		0.000	23:13	
Subject Sample	# 1		23:13	Subject Test Refused
Breath Volume		0.000	23:13	
Air Blank		0.000	23:14	
Control Test		0.081	23:14	
Air Blank		0.000	23:15	
Diagnostics Check		OK	23:15	

Data Download Date/Time: 02/03/2017 00:09

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRIFFIN,KYLE,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: CURRY,RICHARD
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG622905
DRY GAS EXPIRY DATE: 08/16/2018
VIOLATION CODE: DUI
ARREST TIME: 19:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:07

SUBJECT'S LAST NAME: BEASLEY

DATE/TIME OF TEST: 01/30/2017 20:34:04

Diagnostics Check		OK	20:27
Air Blank		0.000	20:27
Control Test		0.083	20:28
Air Blank		0.000	20:28
Subject Sample	# 1	0.128	20:29
Breath Volume		1.773	20:29
Air Blank		0.000	20:29
Air Blank		0.000	20:31
Subject Sample	# 2	0.125	20:32
Breath Volume		2.160	20:32
Air Blank		0.000	20:33
Control Test		0.082	20:33
Air Blank		0.000	20:33
Diagnostics Check		OK	20:34

Data Download Date/Time: 02/03/2017 00:09

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HORNING,PETER,T
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: SKALKO,RYAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG622905
DRY GAS EXPIRY DATE: 08/16/2018
VIOLATION CODE: DUI
ARREST TIME: 01:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:54

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 01/31/2017 02:20:47

Diagnostics Check		OK	02:14
Air Blank		0.000	02:14
Control Test		0.083	02:14
Air Blank		0.000	02:15
Subject Sample	# 1	0.217	02:16
Breath Volume		2.058	02:16
Air Blank		0.000	02:16
Air Blank		0.000	02:18
Subject Sample	# 2	0.210	02:19
Breath Volume		1.812	02:19
Air Blank		0.000	02:19
Control Test		0.082	02:20
Air Blank		0.000	02:20
Diagnostics Check		OK	02:20

Data Download Date/Time: 02/03/2017 00:09

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOWATT,NATHAN,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MOWATT,NATHAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG622905
DRY GAS EXPIRY DATE: 08/16/2018
VIOLATION CODE: DUI
ARREST TIME: 14:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:52

SUBJECT'S LAST NAME: KEELAN

DATE/TIME OF TEST: 01/31/2017 15:21:50

Diagnostics Check		OK	15:13	
Air Blank		0.000	15:13	
Control Test		0.082	15:14	
Air Blank		0.000	15:14	
Subject Sample	# 1		15:17	No Sample Provided
Breath Volume		0.000	15:17	
Air Blank		0.000	15:18	
Air Blank		0.000	15:20	
Subject Sample	# 2		15:20	Subject Test Refused
Breath Volume		0.000	15:20	
Air Blank		0.000	15:20	
Control Test		0.082	15:21	
Air Blank		0.000	15:21	
Diagnostics Check		OK	15:21	

Data Download Date/Time: 02/03/2017 00:09

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005337

INSTRUMENT REGISTERED TO: PERRY POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLEMENT,JOSEPH,E
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: SWINDLE,C
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 12/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG619605
DRY GAS EXPIRY DATE: 07/14/2018
VIOLATION CODE: DUI
ARREST TIME: 16:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:15

SUBJECT'S LAST NAME: TUCKER

DATE/TIME OF TEST: 01/13/2017 16:54:58

Diagnostics Check		OK	16:43	
Air Blank		0.000	16:44	
Control Test		0.080	16:44	
Air Blank		0.000	16:44	
Subject Sample	# 1	0.000	16:48	Volume Not Met
Breath Volume		0.289	16:48	
Air Blank		0.000	16:48	
Air Blank		0.000	16:50	
Subject Sample	# 2	0.000	16:53	Volume Not Met
Breath Volume		0.320	16:53	
Air Blank		0.000	16:54	
Control Test		0.081	16:54	
Air Blank		0.000	16:54	
Diagnostics Check		OK	16:54	

Data Download Date/Time: 01/27/2017 14:41

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005337

INSTRUMENT REGISTERED TO: PERRY POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SWINDLE,CHARLES,M
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: SWINDLE,CHARLES
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 12/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG619605
DRY GAS EXPIRY DATE: 07/14/2018
VIOLATION CODE: DUI
ARREST TIME: 14:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:40

SUBJECT'S LAST NAME: PAINTER

DATE/TIME OF TEST: 01/18/2017 16:07:13

Diagnostics Check		OK	16:00
Air Blank		0.000	16:01
Control Test		0.082	16:01
Air Blank		0.000	16:02
Subject Sample	# 1	0.000	16:02
Breath Volume		1.429	16:02
Air Blank		0.000	16:03
Air Blank		0.000	16:05
Subject Sample	# 2	0.000	16:05
Breath Volume		1.894	16:05
Air Blank		0.000	16:06
Control Test		0.081	16:06
Air Blank		0.000	16:07
Diagnostics Check		OK	16:07

Data Download Date/Time: 01/27/2017 14:41

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005337

INSTRUMENT REGISTERED TO: PERRY POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLEMENT,JOSEPH,E
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: SWINDLE,C
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 12/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG619605
DRY GAS EXPIRY DATE: 07/14/2018
VIOLATION CODE: DUI
ARREST TIME: 07:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:45

SUBJECT'S LAST NAME: GLASS

DATE/TIME OF TEST: 01/27/2017 08:12:21

Diagnostics Check		OK	08:06
Air Blank		0.000	08:06
Control Test		0.082	08:07
Air Blank		0.000	08:07
Subject Sample	# 1	0.000	08:08
Breath Volume		2.597	08:08
Air Blank		0.000	08:08
Air Blank		0.000	08:10
Subject Sample	# 2	0.000	08:11
Breath Volume		2.152	08:11
Air Blank		0.000	08:11
Control Test		0.082	08:11
Air Blank		0.000	08:12
Diagnostics Check		OK	08:12

Data Download Date/Time: 01/27/2017 14:41

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COX,LENARD,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: WEISKOPF,MICHAEL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A2
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:56

SUBJECT'S LAST NAME: MANNERS

DATE/TIME OF TEST: 01/01/2017 02:25:47

Diagnostics Check		OK	02:18
Air Blank		0.000	02:19
Control Test		0.080	02:19
Air Blank		0.000	02:20
Subject Sample	# 1	0.155	02:20
Breath Volume		2.929	02:20
Air Blank		0.000	02:21
Air Blank		0.000	02:23
Subject Sample	# 2	0.146	02:24
Breath Volume		2.507	02:24
Air Blank		0.000	02:24
Control Test		0.079	02:25
Air Blank		0.000	02:25
Diagnostics Check		OK	02:25

Data Download Date/Time: 01/14/2017 19:58

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLI,ANTHONY,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: COX,LENARD
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A2
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:26

SUBJECT'S LAST NAME: PORTU

DATE/TIME OF TEST: 01/05/2017 03:08:10

Diagnostics Check		OK	02:59	
Air Blank		0.000	02:59	
Control Test		0.080	03:00	
Air Blank		0.000	03:00	
Subject Sample	# 1		03:03	No Sample Provided
Breath Volume		0.000	03:03	
Air Blank		0.000	03:04	
Air Blank		0.000	03:06	
Subject Sample	# 2		03:06	Subject Test Refused
Breath Volume		0.000	03:06	
Air Blank		0.000	03:06	
Control Test		0.080	03:07	
Air Blank		0.000	03:07	
Diagnostics Check		Fail	03:07	Voltage/Current Test
Air Blank			03:08	RFI Detect

Data Download Date/Time: 01/14/2017 19:58

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLI,ANTHONY,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: COX,LENARD
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A2
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:26

SUBJECT'S LAST NAME: PORTU

DATE/TIME OF TEST: 01/05/2017 03:14:00

Diagnostics Check		OK	03:11	
Air Blank		0.000	03:11	
Control Test		0.080	03:11	
Air Blank		0.000	03:12	
Subject Sample	# 1		03:12	Subject Test Refused
Breath Volume		0.000	03:12	
Air Blank		0.000	03:13	
Control Test		0.079	03:13	
Air Blank		0.000	03:13	
Diagnostics Check		OK	03:14	

Data Download Date/Time: 01/14/2017 19:58

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COX,LENARD,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: KARAYIANES,MICHAEL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A2
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:47

SUBJECT'S LAST NAME: FOLWELL

DATE/TIME OF TEST: 01/05/2017 22:09:06

Diagnostics Check	Fail	22:08	Voltage/Current Test
Air Blank	0.000	22:09	

Data Download Date/Time: 01/14/2017 19:58

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COX,LENARD,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: KARAYIANES,MICHAEL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A2
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:47

SUBJECT'S LAST NAME: FOLWELL

DATE/TIME OF TEST: 01/05/2017 22:13:17

Diagnostics Check		OK	22:10	
Air Blank		0.000	22:10	
Control Test		0.080	22:11	
Air Blank		0.000	22:11	
Subject Sample	# 1		22:11	Subject Test Refused
Breath Volume		0.000	22:11	
Air Blank		0.000	22:12	
Control Test		0.081	22:12	
Air Blank		0.000	22:13	
Diagnostics Check		OK	22:13	

Data Download Date/Time: 01/14/2017 19:58

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COX,LENARD,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: KARAYIANES,MICHAEL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A2
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:38

SUBJECT'S LAST NAME: PAVLOVIC

DATE/TIME OF TEST: 01/07/2017 00:20:32

Diagnostics Check		OK	00:14	
Air Blank		0.000	00:14	
Control Test		0.078	00:14	
Air Blank		0.000	00:15	
Subject Sample	# 1	0.160	00:15	
Breath Volume		1.171	00:15	
Air Blank		0.000	00:16	
Air Blank		0.000	00:18	
Subject Sample	# 2	0.158	00:18	
Breath Volume		3.304	00:18	
Air Blank		0.000	00:19	
Control Test		0.078	00:19	
Air Blank		0.000	00:20	
Diagnostics Check		Fail	00:20	Voltage/Current Test
Air Blank		0.000	00:20	

Data Download Date/Time: 01/14/2017 19:58

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COX',LENARD,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: KARAYIANES,MICHAEL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A2
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:38

SUBJECT'S LAST NAME: PAVLOVIC

DATE/TIME OF TEST: 01/07/2017 00:28:53

Diagnostics Check		OK	00:22
Air Blank		0.000	00:22
Control Test		0.077	00:22
Air Blank		0.000	00:23
Subject Sample	# 1	0.154	00:24
Breath Volume		3.613	00:24
Air Blank		0.000	00:24
Air Blank		0.000	00:26
Subject Sample	# 2	0.164	00:27
Breath Volume		3.789	00:27
Air Blank		0.000	00:27
Control Test		0.077	00:28
Air Blank		0.000	00:28
Diagnostics Check		OK	00:28

Data Download Date/Time: 01/14/2017 19:58

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COX,LENARD,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: WALDO,AARON
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A2
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: HAYES

DATE/TIME OF TEST: 01/14/2017 01:07:55

Diagnostics Check		OK	01:04	
Air Blank		0.000	01:05	
Control Test		0.079	01:05	
Air Blank		0.000	01:06	
Subject Sample	# 1		01:06	Subject Test Refused
Breath Volume		0.000	01:06	
Air Blank		0.000	01:06	
Control Test		0.079	01:07	
Air Blank		0.000	01:07	
Diagnostics Check		OK	01:07	

Data Download Date/Time: 01/14/2017 19:58

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COX,LENARD,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: WEISKOPF,MICHAEL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 01/14/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A2
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: BRIDGE

DATE/TIME OF TEST: 01/15/2017 02:53:47

Diagnostics Check		OK	02:50	
Air Blank		0.000	02:51	
Control Test		0.080	02:51	
Air Blank		0.000	02:52	
Subject Sample	# 1		02:52	Subject Test Refused
Breath Volume		0.000	02:52	
Air Blank		0.000	02:52	
Control Test		0.079	02:53	
Air Blank		0.000	02:53	
Diagnostics Check		OK	02:53	

Data Download Date/Time: 02/08/2017 21:20

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COX,LENARD,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: KARAYIANES,MICHAEL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 01/14/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A2
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:13

SUBJECT'S LAST NAME: GOTTMAN

DATE/TIME OF TEST: 01/18/2017 23:45:07

Diagnostics Check		OK	23:38
Air Blank		0.000	23:39
Control Test		0.079	23:39
Air Blank		0.000	23:40
Subject Sample	# 1	0.231	23:40
Breath Volume		2.261	23:40
Air Blank		0.000	23:41
Air Blank		0.000	23:43
Subject Sample	# 2	0.228	23:43
Breath Volume		2.527	23:43
Air Blank		0.000	23:44
Control Test		0.079	23:44
Air Blank		0.000	23:45
Diagnostics Check		OK	23:45

Data Download Date/Time: 02/08/2017 21:20

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COX,LENARD,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: WALDO,AARON
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 01/14/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A2
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: GARRIGAN

DATE/TIME OF TEST: 01/19/2017 02:04:15

Diagnostics Check		OK	02:01	
Air Blank		0.000	02:01	
Control Test		0.079	02:01	
Air Blank		0.000	02:02	
Subject Sample	# 1		02:02	Subject Test Refused
Breath Volume		0.000	02:02	
Air Blank		0.000	02:03	
Control Test		0.079	02:03	
Air Blank		0.000	02:04	
Diagnostics Check		OK	02:04	

Data Download Date/Time: 02/08/2017 21:20

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLI,ANTHONY,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: COX,LENARD
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 01/14/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A2
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:42

SUBJECT'S LAST NAME: DOUGHERTY

DATE/TIME OF TEST: 01/28/2017 02:26:49

Diagnostics Check		OK	02:19
Air Blank		0.000	02:19
Control Test		0.081	02:19
Air Blank		0.000	02:20
Subject Sample	# 1	0.138	02:22
Breath Volume		2.132	02:22
Air Blank		0.000	02:22
Air Blank		0.000	02:24
Subject Sample	# 2	0.134	02:25
Breath Volume		1.410	02:25
Air Blank		0.000	02:25
Control Test		0.079	02:26
Air Blank		0.000	02:26
Diagnostics Check		OK	02:26

Data Download Date/Time: 02/08/2017 21:20

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COX,LENARD,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ALLI,ANTHONY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 01/14/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A2
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:22

SUBJECT'S LAST NAME: STEBING

DATE/TIME OF TEST: 01/28/2017 22:51:33

Diagnostics Check		OK	22:45
Air Blank		0.000	22:45
Control Test		0.080	22:46
Air Blank		0.000	22:46
Subject Sample	# 1	0.059	22:46
Breath Volume		2.722	22:46
Air Blank		0.000	22:47
Air Blank		0.000	22:49
Subject Sample	# 2	0.058	22:50
Breath Volume		3.121	22:50
Air Blank		0.000	22:50
Control Test		0.082	22:50
Air Blank		0.000	22:51
Diagnostics Check		OK	22:51

Data Download Date/Time: 02/08/2017 21:20

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COX,LENARD,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ALLI,ANTHONY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 01/14/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A2
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:56

SUBJECT'S LAST NAME: GLYNN JR

DATE/TIME OF TEST: 01/29/2017 03:25:03

Diagnostics Check		OK	03:22	
Air Blank		0.000	03:22	
Control Test		0.080	03:22	
Air Blank		0.000	03:23	
Subject Sample	# 1		03:23	Subject Test Refused
Breath Volume		0.000	03:23	
Air Blank		0.000	03:24	
Control Test		0.080	03:24	
Air Blank		0.000	03:24	
Diagnostics Check		OK	03:25	

Data Download Date/Time: 02/08/2017 21:21

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTER,MATTHEW,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: SAIA,MICHAEL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 01/14/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A2
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: SIMS

DATE/TIME OF TEST: 01/30/2017 00:12:07

Diagnostics Check		OK	00:05
Air Blank		0.000	00:06
Control Test		0.079	00:06
Air Blank		0.000	00:07
Subject Sample	# 1	0.156	00:07
Breath Volume		2.269	00:07
Air Blank		0.000	00:08
Air Blank		0.000	00:10
Subject Sample	# 2	0.154	00:10
Breath Volume		1.843	00:10
Air Blank		0.000	00:11
Control Test		0.080	00:11
Air Blank		0.000	00:12
Diagnostics Check		OK	00:12

Data Download Date/Time: 02/08/2017 21:21

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CALLAHAN,WESLEY,S
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: INTRAVICHIT,SOMSAK
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 01/14/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A2
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 08:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:13

SUBJECT'S LAST NAME: P636421902920

DATE/TIME OF TEST: 01/30/2017 09:53:30

Diagnostics Check		OK	09:47
Air Blank		0.000	09:47
Control Test		0.078	09:47
Air Blank		0.000	09:48
Subject Sample	# 1	0.124	09:48
Breath Volume		3.945	09:48
Air Blank		0.000	09:49
Air Blank		0.000	09:51
Subject Sample	# 2	0.126	09:51
Breath Volume		3.691	09:51
Air Blank		0.000	09:52
Control Test		0.078	09:52
Air Blank		0.000	09:53
Diagnostics Check		OK	09:53

Data Download Date/Time: 02/08/2017 21:21

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005402

INSTRUMENT REGISTERED TO: PORT ORANGE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MANDESE,ANGELO,
BREATH TEST OPERATOR AGENCY: Port Orange PD
ARRESTING OFFICER: FLEMING,BERNARD
ARRESTING OFFICER AGENCY: Port Orange PD
LAST AGENCY INSPECTION: 12/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1836890
DRY GAS EXPIRY DATE: 03/26/2018
VIOLATION CODE: DUI
ARREST TIME: 01:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: MITCHELL

DATE/TIME OF TEST: 01/05/2017 03:31:18

Diagnostics Check		OK	03:25
Air Blank		0.000	03:25
Control Test		0.082	03:26
Air Blank		0.000	03:26
Subject Sample	# 1	0.000	03:27
Breath Volume		2.312	03:27
Air Blank		0.000	03:27
Air Blank		0.000	03:29
Subject Sample	# 2	0.000	03:29
Breath Volume		2.757	03:29
Air Blank		0.000	03:30
Control Test		0.081	03:30
Air Blank		0.000	03:31
Diagnostics Check		OK	03:31

Data Download Date/Time: 01/20/2017 14:39

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005402

INSTRUMENT REGISTERED TO: PORT ORANGE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NOLAN,JONATHAN,W
BREATH TEST OPERATOR AGENCY: Port Orange PD
ARRESTING OFFICER: NOLAN,JONATHAN
ARRESTING OFFICER AGENCY: Port Orange PD
LAST AGENCY INSPECTION: 12/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1836890
DRY GAS EXPIRY DATE: 03/26/2018
VIOLATION CODE: DUI
ARREST TIME: 01:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: ALFORD

DATE/TIME OF TEST: 01/07/2017 02:42:23

Diagnostics Check		OK	02:35
Air Blank		0.000	02:35
Control Test		0.082	02:36
Air Blank		0.000	02:36
Subject Sample	# 1	0.202	02:37
Breath Volume		1.292	02:37
Air Blank		0.000	02:37
Air Blank		0.000	02:39
Subject Sample	# 2	0.191	02:40
Breath Volume		1.554	02:40
Air Blank		0.000	02:41
Control Test		0.081	02:41
Air Blank		0.000	02:42
Diagnostics Check		OK	02:42

Data Download Date/Time: 01/20/2017 14:39

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005402

INSTRUMENT REGISTERED TO: PORT ORANGE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MANDESE,ANGELO,
BREATH TEST OPERATOR AGENCY: Port Orange PD
ARRESTING OFFICER: MANDESE,ANGELO
ARRESTING OFFICER AGENCY: Port Orange PD
LAST AGENCY INSPECTION: 12/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1836890
DRY GAS EXPIRY DATE: 03/26/2018
VIOLATION CODE: ADMIN
ARREST TIME: 18:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:30

SUBJECT'S LAST NAME: RAINES

DATE/TIME OF TEST: 01/13/2017 18:59:23

Diagnostics Check		OK	18:53
Air Blank		0.000	18:53
Control Test		0.083	18:54
Air Blank		0.000	18:54
Subject Sample	# 1	0.000	18:55
Breath Volume		3.707	18:55
Air Blank		0.000	18:55
Air Blank		0.000	18:57
Subject Sample	# 2	0.000	18:58
Breath Volume		3.730	18:58
Air Blank		0.000	18:58
Control Test		0.081	18:58
Air Blank		0.000	18:59
Diagnostics Check		OK	18:59

Data Download Date/Time: 01/20/2017 14:39

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005402

INSTRUMENT REGISTERED TO: PORT ORANGE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAINES,CHRISTOPHER,G
BREATH TEST OPERATOR AGENCY: Port Orange PD
ARRESTING OFFICER: RAINES,CHRISTOPHER
ARRESTING OFFICER AGENCY: Port Orange PD
LAST AGENCY INSPECTION: 12/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1836890
DRY GAS EXPIRY DATE: 03/26/2018
VIOLATION CODE: ADMIN
ARREST TIME: 18:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:30

SUBJECT'S LAST NAME: MANDESE

DATE/TIME OF TEST: 01/13/2017 19:07:36

Diagnostics Check		OK	19:01
Air Blank		0.000	19:02
Control Test		0.082	19:02
Air Blank		0.000	19:02
Subject Sample	# 1	0.000	19:03
Breath Volume		3.042	19:03
Air Blank		0.000	19:03
Air Blank		0.000	19:05
Subject Sample	# 2	0.000	19:06
Breath Volume		3.445	19:06
Air Blank		0.000	19:06
Control Test		0.082	19:07
Air Blank		0.000	19:07
Diagnostics Check		OK	19:07

Data Download Date/Time: 01/20/2017 14:39

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005402

INSTRUMENT REGISTERED TO: PORT ORANGE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MANDESE,ANGELO,
BREATH TEST OPERATOR AGENCY: Port Orange PD
ARRESTING OFFICER: MANDESE,ANGELO
ARRESTING OFFICER AGENCY: Port Orange PD
LAST AGENCY INSPECTION: 12/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1836890
DRY GAS EXPIRY DATE: 03/26/2018
VIOLATION CODE: ADMIN
ARREST TIME: 21:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:10

SUBJECT'S LAST NAME: SANTANA

DATE/TIME OF TEST: 01/13/2017 21:42:49

Diagnostics Check		OK	21:36
Air Blank		0.000	21:37
Control Test		0.081	21:37
Air Blank		0.000	21:38
Subject Sample	# 1	0.000	21:38
Breath Volume		2.812	21:38
Air Blank		0.000	21:39
Air Blank		0.000	21:41
Subject Sample	# 2	0.000	21:41
Breath Volume		3.023	21:41
Air Blank		0.000	21:41
Control Test		0.081	21:42
Air Blank		0.000	21:42
Diagnostics Check		OK	21:42

Data Download Date/Time: 01/20/2017 14:39

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005402

INSTRUMENT REGISTERED TO: PORT ORANGE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BENEZETTE,BENJAMIN,J
BREATH TEST OPERATOR AGENCY: Port Orange PD
ARRESTING OFFICER: SEGURA,SAMUEL
ARRESTING OFFICER AGENCY: Port Orange PD
LAST AGENCY INSPECTION: 12/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1836890
DRY GAS EXPIRY DATE: 03/26/2018
VIOLATION CODE: DUI
ARREST TIME: 23:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: HARPER

DATE/TIME OF TEST: 01/15/2017 00:38:08

Diagnostics Check		OK	00:31
Air Blank		0.000	00:31
Control Test		0.081	00:32
Air Blank		0.000	00:32
Subject Sample	# 1	0.137	00:33
Breath Volume		2.789	00:33
Air Blank		0.000	00:34
Air Blank		0.000	00:36
Subject Sample	# 2	0.136	00:36
Breath Volume		2.121	00:36
Air Blank		0.000	00:37
Control Test		0.080	00:37
Air Blank		0.000	00:38
Diagnostics Check		OK	00:38

Data Download Date/Time: 01/20/2017 14:39

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005402

INSTRUMENT REGISTERED TO: PORT ORANGE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BENEZETTE,BENJAMIN,J
BREATH TEST OPERATOR AGENCY: Port Orange PD
ARRESTING OFFICER: MCLEAN,JESSICA
ARRESTING OFFICER AGENCY: Port Orange PD
LAST AGENCY INSPECTION: 12/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1836890
DRY GAS EXPIRY DATE: 03/26/2018
VIOLATION CODE: DUI
ARREST TIME: 20:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:41

SUBJECT'S LAST NAME: CONREY

DATE/TIME OF TEST: 01/15/2017 21:10:43

Diagnostics Check		OK	21:04
Air Blank		0.000	21:04
Control Test		0.082	21:05
Air Blank		0.000	21:05
Subject Sample	# 1	0.201	21:06
Breath Volume		3.417	21:06
Air Blank		0.000	21:06
Air Blank		0.000	21:08
Subject Sample	# 2	0.206	21:09
Breath Volume		2.613	21:09
Air Blank		0.000	21:09
Control Test		0.081	21:10
Air Blank		0.000	21:10
Diagnostics Check		OK	21:10

Data Download Date/Time: 01/20/2017 14:39

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005402

INSTRUMENT REGISTERED TO: PORT ORANGE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FIELDS,ERIC,R
BREATH TEST OPERATOR AGENCY: Port Orange PD
ARRESTING OFFICER: FIELDS,ERIC
ARRESTING OFFICER AGENCY: Port Orange PD
LAST AGENCY INSPECTION: 12/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1836890
DRY GAS EXPIRY DATE: 03/26/2018
VIOLATION CODE: ADMIN
ARREST TIME: 13:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:00

SUBJECT'S LAST NAME: FISHER

DATE/TIME OF TEST: 01/16/2017 15:32:22

Diagnostics Check		OK	15:26
Air Blank		0.000	15:27
Control Test		0.082	15:27
Air Blank		0.000	15:27
Subject Sample	# 1	0.000	15:28
Breath Volume		1.375	15:28
Air Blank		0.000	15:28
Air Blank		0.000	15:30
Subject Sample	# 2	0.000	15:31
Breath Volume		1.425	15:31
Air Blank		0.000	15:31
Control Test		0.082	15:31
Air Blank		0.000	15:32
Diagnostics Check		OK	15:32

Data Download Date/Time: 01/20/2017 14:39

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005402

INSTRUMENT REGISTERED TO: PORT ORANGE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KURTZ,THEODORE,F
BREATH TEST OPERATOR AGENCY: Port Orange PD
ARRESTING OFFICER: ALEVISATOS,JARED
ARRESTING OFFICER AGENCY: Port Orange PD
LAST AGENCY INSPECTION: 12/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1836890
DRY GAS EXPIRY DATE: 03/26/2018
VIOLATION CODE: DUI
ARREST TIME: 19:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:02

SUBJECT'S LAST NAME: CLEMENTS

DATE/TIME OF TEST: 01/16/2017 20:39:09

Diagnostics Check		OK	20:24	
Air Blank		0.000	20:25	
Control Test		0.083	20:25	
Air Blank		0.000	20:25	
Subject Sample	# 1	0.221	20:29	Volume Not Met
Breath Volume		0.437	20:29	
Air Blank		0.000	20:29	
Air Blank		0.000	20:31	
Subject Sample	# 2	0.227	20:33	
Breath Volume		1.910	20:33	
Air Blank		0.000	20:34	
Air Blank		0.000	20:36	
Subject Sample	# 3	0.241	20:37	
Breath Volume		1.207	20:37	
Air Blank		0.000	20:38	
Control Test		0.079	20:38	
Air Blank		0.000	20:39	
Diagnostics Check		OK	20:39	

Data Download Date/Time: 01/20/2017 14:39

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005402

INSTRUMENT REGISTERED TO: PORT ORANGE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCKAY,ANDREW,G
BREATH TEST OPERATOR AGENCY: Port Orange PD
ARRESTING OFFICER: MCKAY,ANDY
ARRESTING OFFICER AGENCY: Port Orange PD
LAST AGENCY INSPECTION: 12/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1836890
DRY GAS EXPIRY DATE: 03/26/2018
VIOLATION CODE: ADMIN
ARREST TIME: 18:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:30

SUBJECT'S LAST NAME: SCACCIA

DATE/TIME OF TEST: 01/17/2017 19:16:42

Diagnostics Check		OK	19:11
Air Blank		0.000	19:11
Control Test		0.081	19:11
Air Blank		0.000	19:12
Subject Sample	# 1	0.000	19:12
Breath Volume		2.621	19:12
Air Blank		0.000	19:13
Air Blank		0.000	19:15
Subject Sample	# 2	0.000	19:15
Breath Volume		2.406	19:15
Air Blank		0.000	19:15
Control Test		0.081	19:16
Air Blank		0.000	19:16
Diagnostics Check		OK	19:16

Data Download Date/Time: 01/20/2017 14:39

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005402

INSTRUMENT REGISTERED TO: PORT ORANGE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MANDESE,ANGELO,
BREATH TEST OPERATOR AGENCY: Port Orange PD
ARRESTING OFFICER: MCLEAN,JESSICA
ARRESTING OFFICER AGENCY: Port Orange PD
LAST AGENCY INSPECTION: 12/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1836890
DRY GAS EXPIRY DATE: 03/26/2018
VIOLATION CODE: DUI
ARREST TIME: 01:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:07

SUBJECT'S LAST NAME: O NEAL

DATE/TIME OF TEST: 01/20/2017 02:39:23

Diagnostics Check		OK	02:32
Air Blank		0.000	02:33
Control Test		0.080	02:33
Air Blank		0.000	02:34
Subject Sample	# 1	0.237	02:34
Breath Volume		2.351	02:34
Air Blank		0.000	02:35
Air Blank		0.000	02:37
Subject Sample	# 2	0.237	02:37
Breath Volume		2.679	02:37
Air Blank		0.000	02:38
Control Test		0.081	02:38
Air Blank		0.000	02:39
Diagnostics Check		OK	02:39

Data Download Date/Time: 01/20/2017 14:39

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005402

INSTRUMENT REGISTERED TO: PORT ORANGE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KURTZ,THEODORE,F
BREATH TEST OPERATOR AGENCY: Port Orange PD
ARRESTING OFFICER: KURTZ,THEODORE
ARRESTING OFFICER AGENCY: Port Orange PD
LAST AGENCY INSPECTION: 01/20/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1836890
DRY GAS EXPIRY DATE: 03/26/2018
VIOLATION CODE: OTHER
ARREST TIME: 04:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:40

SUBJECT'S LAST NAME: VELEZ

DATE/TIME OF TEST: 01/23/2017 05:16:34

Diagnostics Check		OK	05:10
Air Blank		0.000	05:11
Control Test		0.081	05:11
Air Blank		0.000	05:11
Subject Sample	# 1	0.000	05:12
Breath Volume		2.882	05:12
Air Blank		0.000	05:12
Air Blank		0.000	05:14
Subject Sample	# 2	0.000	05:15
Breath Volume		2.785	05:15
Air Blank		0.000	05:15
Control Test		0.082	05:15
Air Blank		0.000	05:16
Diagnostics Check		OK	05:16

Data Download Date/Time: 02/26/2017 13:20

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005402

INSTRUMENT REGISTERED TO: PORT ORANGE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERSON,JESSE,C
BREATH TEST OPERATOR AGENCY: Port Orange PD
ARRESTING OFFICER: PIERSON,JESSE
ARRESTING OFFICER AGENCY: Port Orange PD
LAST AGENCY INSPECTION: 01/20/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1836890
DRY GAS EXPIRY DATE: 03/26/2018
VIOLATION CODE: ADMIN
ARREST TIME: 14:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:30

SUBJECT'S LAST NAME: SLEASE

DATE/TIME OF TEST: 01/24/2017 15:09:46

Diagnostics Check		OK	15:03
Air Blank		0.000	15:04
Control Test		0.082	15:04
Air Blank		0.000	15:05
Subject Sample	# 1	0.000	15:05
Breath Volume		2.882	15:05
Air Blank		0.000	15:06
Air Blank		0.000	15:08
Subject Sample	# 2	0.000	15:08
Breath Volume		3.046	15:08
Air Blank		0.000	15:08
Control Test		0.082	15:09
Air Blank		0.000	15:09
Diagnostics Check		OK	15:09

Data Download Date/Time: 02/26/2017 13:20

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005537

INSTRUMENT REGISTERED TO: DAYTONA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIMA,MICHAEL,A
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: TARR,BRANDON
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 12/19/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14968
DRY GAS EXPIRY DATE: 08/06/2018
VIOLATION CODE: DUI
ARREST TIME: 19:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:44

SUBJECT'S LAST NAME: STEWART

DATE/TIME OF TEST: 01/02/2017 20:18:34

Diagnostics Check		OK	20:12
Air Blank		0.000	20:12
Control Test		0.078	20:13
Air Blank		0.000	20:13
Subject Sample	# 1	0.122	20:14
Breath Volume		3.472	20:14
Air Blank		0.000	20:14
Air Blank		0.000	20:16
Subject Sample	# 2	0.120	20:16
Breath Volume		3.195	20:16
Air Blank		0.000	20:17
Control Test		0.077	20:17
Air Blank		0.000	20:18
Diagnostics Check		OK	20:18

Data Download Date/Time: 01/26/2017 14:33

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005537

INSTRUMENT REGISTERED TO: DAYTONA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CONLEY,SHAWNA,A
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: CONLEY,SHAWNA
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 12/19/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14968
DRY GAS EXPIRY DATE: 08/06/2018
VIOLATION CODE: DUI
ARREST TIME: 18:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:06

SUBJECT'S LAST NAME: FARMER

DATE/TIME OF TEST: 01/13/2017 19:38:56

Diagnostics Check		OK	19:27	
Air Blank		0.000	19:27	
Control Test		0.080	19:27	
Air Blank		0.000	19:28	
Subject Sample	# 1	0.230	19:31	Volume Not Met
Breath Volume		0.097	19:31	
Air Blank		0.000	19:32	
Air Blank		0.000	19:34	
Subject Sample	# 2	0.214	19:37	Volume Not Met
Breath Volume		0.164	19:37	
Air Blank		0.000	19:37	
Control Test		0.077	19:38	
Air Blank		0.000	19:38	
Diagnostics Check		OK	19:38	

Data Download Date/Time: 01/26/2017 14:33

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005537

INSTRUMENT REGISTERED TO: DAYTONA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARD,JAMES,A
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: WARD,JAMES
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 12/19/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14968
DRY GAS EXPIRY DATE: 08/06/2018
VIOLATION CODE: DUI
ARREST TIME: 20:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:26

SUBJECT'S LAST NAME: COVINGTON

DATE/TIME OF TEST: 01/17/2017 20:58:59

Diagnostics Check		OK	20:52
Air Blank		0.000	20:52
Control Test		0.080	20:52
Air Blank		0.000	20:53
Subject Sample	# 1	0.254	20:54
Breath Volume		3.066	20:54
Air Blank		0.000	20:54
Air Blank		0.000	20:56
Subject Sample	# 2	0.247	20:57
Breath Volume		3.195	20:57
Air Blank		0.000	20:58
Control Test		0.079	20:58
Air Blank		0.000	20:58
Diagnostics Check		OK	20:58

Data Download Date/Time: 01/26/2017 14:33

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005537

INSTRUMENT REGISTERED TO: DAYTONA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CONLEY,SHAWNA,A
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: MAHOI,MUSA
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 12/19/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14968
DRY GAS EXPIRY DATE: 08/06/2018
VIOLATION CODE: DUI
ARREST TIME: 21:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: ALMEIDA

DATE/TIME OF TEST: 01/19/2017 00:44:22

Diagnostics Check		OK	00:41	
Air Blank		0.000	00:41	
Control Test		0.079	00:42	
Air Blank		0.000	00:42	
Subject Sample	# 1		00:42	Subject Test Refused
Breath Volume		0.000	00:42	
Air Blank		0.000	00:43	
Control Test		0.079	00:43	
Air Blank		0.000	00:44	
Diagnostics Check		OK	00:44	

Data Download Date/Time: 01/26/2017 14:33

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005537

INSTRUMENT REGISTERED TO: DAYTONA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARD,JAMES,A
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: WARD,JAMES
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 12/19/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14968
DRY GAS EXPIRY DATE: 08/06/2018
VIOLATION CODE: DUI
ARREST TIME: 00:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:22

SUBJECT'S LAST NAME: WELLS

DATE/TIME OF TEST: 01/23/2017 00:49:12

Diagnostics Check		OK	00:44	
Air Blank		0.000	00:45	
Control Test		0.079	00:45	
Air Blank		0.000	00:46	
Subject Sample	# 1		00:47	Subject Test Refused
Breath Volume		0.468	00:47	
Air Blank		0.000	00:48	
Control Test		0.079	00:48	
Air Blank		0.000	00:49	
Diagnostics Check		OK	00:49	

Data Download Date/Time: 01/26/2017 14:33

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005537

INSTRUMENT REGISTERED TO: DAYTONA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARD,JAMES,A
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: WARD,JAMES
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 12/19/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14968
DRY GAS EXPIRY DATE: 08/06/2018
VIOLATION CODE: DUI
ARREST TIME: 00:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:22

SUBJECT'S LAST NAME: WELLS

DATE/TIME OF TEST: 01/23/2017 00:57:58

Diagnostics Check		OK	00:50
Air Blank		0.000	00:50
Control Test		0.078	00:51
Air Blank		0.000	00:51
Subject Sample	# 1	0.154	00:52
Breath Volume		1.585	00:52
Air Blank		0.000	00:53
Air Blank		0.000	00:55
Subject Sample	# 2	0.138	00:56
Breath Volume		1.890	00:56
Air Blank		0.000	00:57
Control Test		0.078	00:57
Air Blank		0.000	00:57
Diagnostics Check		OK	00:57

Data Download Date/Time: 01/26/2017 14:33

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005537

INSTRUMENT REGISTERED TO: DAYTONA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CONLEY,SHAWNA,A
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: EDWARDS,LANDON
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 01/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14968
DRY GAS EXPIRY DATE: 08/06/2018
VIOLATION CODE: DUI
ARREST TIME: 02:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:23

SUBJECT'S LAST NAME: MANRRIQUE CORTEZ

DATE/TIME OF TEST: 01/29/2017 03:52:31

Diagnostics Check		OK	03:45
Air Blank		0.000	03:45
Control Test		0.080	03:45
Air Blank		0.000	03:46
Subject Sample	# 1	0.220	03:47
Breath Volume		3.285	03:47
Air Blank		0.000	03:48
Air Blank		0.000	03:49
Subject Sample	# 2	0.222	03:50
Breath Volume		3.105	03:50
Air Blank		0.000	03:51
Control Test		0.079	03:51
Air Blank		0.000	03:52
Diagnostics Check		OK	03:52

Data Download Date/Time: 02/13/2017 14:19

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASTILLO,CHRISTIAN,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: FERRO,LISA
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 01/04/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 17:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:30

SUBJECT'S LAST NAME: DAVILA

DATE/TIME OF TEST: 01/06/2017 17:58:07

Diagnostics Check		OK	17:50
Air Blank		0.000	17:51
Control Test		0.079	17:51
Air Blank		0.000	17:51
Subject Sample	# 1	0.201	17:53
Breath Volume		2.343	17:53
Air Blank		0.000	17:53
Air Blank		0.000	17:55
Subject Sample	# 2	0.198	17:56
Breath Volume		3.914	17:56
Air Blank		0.000	17:57
Control Test		0.079	17:57
Air Blank		0.000	17:58
Diagnostics Check		OK	17:58

Data Download Date/Time: 01/26/2017 12:16

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAYARD,CLIFFORD,P
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: CAYARD,CLIFFORD
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 01/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:57

SUBJECT'S LAST NAME: MORALES

DATE/TIME OF TEST: 01/10/2017 21:21:05

Diagnostics Check		OK	21:18	
Air Blank		0.000	21:18	
Control Test		0.078	21:18	
Air Blank		0.000	21:19	
Subject Sample	# 1		21:19	Subject Test Refused
Breath Volume		0.000	21:19	
Air Blank		0.000	21:20	
Control Test		0.079	21:20	
Air Blank		0.000	21:20	
Diagnostics Check		OK	21:21	

Data Download Date/Time: 01/26/2017 12:16

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 01/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: MERCADO

DATE/TIME OF TEST: 01/20/2017 01:11:31

Diagnostics Check		OK	01:02
Air Blank		0.000	01:02
Control Test		0.079	01:03
Air Blank		0.000	01:03
Subject Sample	# 1	0.000	01:06
Breath Volume		1.402	01:06
Air Blank		0.000	01:06
Air Blank		0.000	01:08
Subject Sample	# 2	0.000	01:10
Breath Volume		1.128	01:10
Air Blank		0.000	01:10
Control Test		0.078	01:10
Air Blank		0.000	01:11
Diagnostics Check		OK	01:11

Data Download Date/Time: 01/26/2017 12:16

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROIG,DAWN,M
BREATH TEST OPERATOR AGENCY: Miami Shores PD
ARRESTING OFFICER: ROIG,DAWN
ARRESTING OFFICER AGENCY: Miami Shores PD
LAST AGENCY INSPECTION: 01/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 16:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:07

SUBJECT'S LAST NAME: MCDOUGALL

DATE/TIME OF TEST: 01/20/2017 17:33:07

Diagnostics Check		OK	17:28	
Air Blank		0.000	17:28	
Control Test		0.077	17:28	
Air Blank		0.000	17:29	
Subject Sample	# 1		17:31	Subject Test Refused
Breath Volume		0.000	17:31	
Air Blank		0.000	17:32	
Control Test		0.077	17:32	
Air Blank		0.000	17:33	
Diagnostics Check		OK	17:33	

Data Download Date/Time: 01/26/2017 12:16

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JANVIER,FRITZ,G
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: JANVIER,FRITZ
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 01/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 05:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:54

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 01/21/2017 06:23:48

Diagnostics Check		OK	06:14
Air Blank		0.000	06:14
Control Test		0.080	06:15
Air Blank		0.000	06:15
Subject Sample	# 1	0.086	06:17
Breath Volume		2.054	06:17
Air Blank		0.000	06:18
Air Blank		0.000	06:20
Subject Sample	# 2	0.088	06:22
Breath Volume		1.699	06:22
Air Blank		0.000	06:22
Control Test		0.079	06:23
Air Blank		0.000	06:23
Diagnostics Check		OK	06:23

Data Download Date/Time: 01/26/2017 12:16

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AGINS,LYLE,P
BREATH TEST OPERATOR AGENCY: Sunny Isles Beach PD
ARRESTING OFFICER: MORENO,T
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 01/04/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:04

SUBJECT'S LAST NAME: TRANA

DATE/TIME OF TEST: 01/22/2017 23:40:13

Diagnostics Check		OK	23:27	
Air Blank		0.000	23:28	
Control Test		0.077	23:28	
Air Blank		0.000	23:29	
Subject Sample	# 1	0.170	23:32	Volume Not Met
Breath Volume		0.332	23:32	
Air Blank		0.000	23:33	
Air Blank		0.000	23:34	
Subject Sample	# 2	0.212	23:35	
Breath Volume		2.011	23:35	
Air Blank		0.000	23:36	
Air Blank		0.000	23:38	
Subject Sample	# 3	0.207	23:38	
Breath Volume		2.046	23:38	
Air Blank		0.000	23:39	
Control Test		0.076	23:39	
Air Blank		0.000	23:40	
Diagnostics Check		OK	23:40	

Data Download Date/Time: 01/26/2017 12:16

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOORE,CORY,J
BREATH TEST OPERATOR AGENCY: Miami Shores PD
ARRESTING OFFICER: MOORE,CORY
ARRESTING OFFICER AGENCY: Miami Shores PD
LAST AGENCY INSPECTION: 12/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:34

SUBJECT'S LAST NAME: VEJE

DATE/TIME OF TEST: 01/02/2017 23:57:27

Diagnostics Check		OK	23:54	
Air Blank		0.000	23:55	
Control Test		0.078	23:55	
Air Blank		0.000	23:55	
Subject Sample	# 1		23:56	Subject Test Refused
Breath Volume		0.000	23:56	
Air Blank		0.000	23:56	
Control Test		0.079	23:56	
Air Blank		0.000	23:57	
Diagnostics Check		OK	23:57	

Data Download Date/Time: 01/20/2017 09:54

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OGDEN,JOHN,L
BREATH TEST OPERATOR AGENCY: Aventura PD
ARRESTING OFFICER: OGDEN,JOHN
ARRESTING OFFICER AGENCY: Aventura PD
LAST AGENCY INSPECTION: 12/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: MORENO DIAZ

DATE/TIME OF TEST: 01/07/2017 02:11:21

Diagnostics Check		OK	02:01	
Air Blank		0.000	02:02	
Control Test		0.076	02:02	
Air Blank		0.000	02:03	
Subject Sample	# 1	0.184	02:06	Volume Not Met
Breath Volume		0.078	02:06	
Air Blank		0.000	02:06	
Air Blank		0.000	02:08	
Subject Sample	# 2		02:09	Subject Test Refused
Breath Volume		0.242	02:09	
Air Blank		0.000	02:10	
Control Test		0.078	02:10	
Air Blank		0.000	02:11	
Diagnostics Check		OK	02:11	

Data Download Date/Time: 01/20/2017 09:55

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AGINS,LYLE,P
BREATH TEST OPERATOR AGENCY: Sunny Isles Beach PD
ARRESTING OFFICER: AGINS,LYLE
ARRESTING OFFICER AGENCY: Sunny Isles Beach PD
LAST AGENCY INSPECTION: 12/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 18:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:22

SUBJECT'S LAST NAME: ABRAHAM

DATE/TIME OF TEST: 01/09/2017 18:49:13

Diagnostics Check		OK	18:42
Air Blank		0.000	18:43
Control Test		0.078	18:43
Air Blank		0.000	18:44
Subject Sample	# 1	0.124	18:44
Breath Volume		3.871	18:44
Air Blank		0.000	18:45
Air Blank		0.000	18:47
Subject Sample	# 2	0.120	18:47
Breath Volume		3.562	18:47
Air Blank		0.000	18:48
Control Test		0.076	18:48
Air Blank		0.000	18:49
Diagnostics Check		OK	18:49

Data Download Date/Time: 01/20/2017 09:55

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PUPO,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Brooker PD
ARRESTING OFFICER: PUPO,MICHAEL
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 12/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: OLSEN

DATE/TIME OF TEST: 01/12/2017 03:07:31

Diagnostics Check		OK	03:00
Air Blank		0.000	03:00
Control Test		0.078	03:01
Air Blank		0.000	03:01
Subject Sample	# 1	0.213	03:02
Breath Volume		2.066	03:02
Air Blank		0.000	03:03
Air Blank		0.000	03:05
Subject Sample	# 2	0.223	03:05
Breath Volume		2.230	03:05
Air Blank		0.000	03:06
Control Test		0.077	03:06
Air Blank		0.000	03:07
Diagnostics Check		OK	03:07

Data Download Date/Time: 01/20/2017 09:55

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JANVIER,FRITZ,G
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: JANVIER,FRITZ
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 12/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: DEERING

DATE/TIME OF TEST: 01/16/2017 01:40:40

Diagnostics Check		OK	01:37	
Air Blank		0.000	01:38	
Control Test		0.079	01:38	
Air Blank		0.000	01:39	
Subject Sample	# 1		01:39	Subject Test Refused
Breath Volume		0.000	01:39	
Air Blank		0.000	01:39	
Control Test		0.079	01:40	
Air Blank		0.000	01:40	
Diagnostics Check		OK	01:40	

Data Download Date/Time: 01/20/2017 09:55

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRANZOSA,NICHOLAS,P
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: STREETZEL,JONATHAN
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 12/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 18:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:28

SUBJECT'S LAST NAME: CHAVES

DATE/TIME OF TEST: 01/16/2017 19:53:05

Diagnostics Check		OK	19:49	
Air Blank		0.000	19:50	
Control Test		0.079	19:50	
Air Blank		0.000	19:50	
Subject Sample	# 1		19:51	Subject Test Refused
Breath Volume		0.000	19:51	
Air Blank		0.000	19:52	
Control Test		0.079	19:52	
Air Blank		0.000	19:53	
Diagnostics Check		OK	19:53	

Data Download Date/Time: 01/20/2017 09:55

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AGINS,LYLE,P
BREATH TEST OPERATOR AGENCY: Sunny Isles Beach PD
ARRESTING OFFICER: AGINS,LYLE
ARRESTING OFFICER AGENCY: Sunny Isles Beach PD
LAST AGENCY INSPECTION: 12/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:51

SUBJECT'S LAST NAME: YEVSEYEV

DATE/TIME OF TEST: 01/18/2017 02:18:52

Diagnostics Check		OK	02:15	
Air Blank		0.000	02:16	
Control Test		0.079	02:16	
Air Blank		0.000	02:17	
Subject Sample	# 1		02:17	Subject Test Refused
Breath Volume		0.000	02:17	
Air Blank		0.000	02:17	
Control Test		0.077	02:18	
Air Blank		0.000	02:18	
Diagnostics Check		OK	02:18	

Data Download Date/Time: 01/20/2017 09:55

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEY,LAZARO,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: PEREZ,EDWIN
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 01/20/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:39

SUBJECT'S LAST NAME: SOSA FERNANDEZ

DATE/TIME OF TEST: 01/22/2017 02:06:15

Diagnostics Check		OK	01:59
Air Blank		0.000	02:00
Control Test		0.079	02:00
Air Blank		0.000	02:00
Subject Sample	# 1	0.112	02:01
Breath Volume		2.343	02:01
Air Blank		0.000	02:02
Air Blank		0.000	02:04
Subject Sample	# 2	0.107	02:04
Breath Volume		2.742	02:04
Air Blank		0.000	02:05
Control Test		0.079	02:05
Air Blank		0.000	02:06
Diagnostics Check		OK	02:06

Data Download Date/Time: 02/14/2017 13:48

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESPINOZA,PABLO,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: FRANCO,GREGORY
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 01/20/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:11

SUBJECT'S LAST NAME: GELDRES AGUILAR

DATE/TIME OF TEST: 01/22/2017 03:44:16

Diagnostics Check		OK	03:34
Air Blank		0.000	03:35
Control Test		0.079	03:35
Air Blank		0.000	03:36
Subject Sample	# 1	0.219	03:36
Breath Volume		2.828	03:36
Air Blank		0.000	03:37
Air Blank		0.000	03:39
Subject Sample	# 2	0.189	03:39
Breath Volume		1.476	03:39
Air Blank		0.000	03:40
Air Blank		0.000	03:42
Subject Sample	# 3	0.223	03:42
Breath Volume		3.480	03:42
Air Blank		0.000	03:43
Control Test		0.078	03:43
Air Blank		0.000	03:44
Diagnostics Check		OK	03:44

Data Download Date/Time: 02/14/2017 13:48

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEY,LAZARO,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: FERRO,LISA
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 01/20/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:51

SUBJECT'S LAST NAME: VAZQUEZ

DATE/TIME OF TEST: 01/22/2017 04:33:05

Diagnostics Check		OK	04:30	
Air Blank		0.000	04:30	
Control Test		0.079	04:31	
Air Blank		0.000	04:31	
Subject Sample	# 1		04:31	Subject Test Refused
Breath Volume		0.000	04:31	
Air Blank		0.000	04:32	
Control Test		0.078	04:32	
Air Blank		0.000	04:33	
Diagnostics Check		OK	04:33	

Data Download Date/Time: 02/14/2017 13:48

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEY,LAZARO,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: FERRO,LISA
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 01/20/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:16

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 01/28/2017 23:52:32

Diagnostics Check		OK	23:49	
Air Blank		0.000	23:50	
Control Test		0.079	23:50	
Air Blank		0.000	23:50	
Subject Sample	# 1		23:51	Subject Test Refused
Breath Volume		0.000	23:51	
Air Blank		0.000	23:51	
Control Test		0.079	23:51	
Air Blank		0.000	23:52	
Diagnostics Check		OK	23:52	

Data Download Date/Time: 02/14/2017 13:48

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEY,LAZARO,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: PEREZ,EDWIN
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 01/20/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:46

SUBJECT'S LAST NAME: DELACT

DATE/TIME OF TEST: 01/29/2017 00:33:13

Diagnostics Check		OK	00:27
Air Blank		0.000	00:27
Control Test		0.079	00:27
Air Blank		0.000	00:28
Subject Sample	# 1	0.152	00:28
Breath Volume		1.632	00:28
Air Blank		0.000	00:29
Air Blank		0.000	00:31
Subject Sample	# 2	0.159	00:31
Breath Volume		1.691	00:31
Air Blank		0.000	00:32
Control Test		0.079	00:32
Air Blank		0.000	00:33
Diagnostics Check		OK	00:33

Data Download Date/Time: 02/14/2017 13:48

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESPINOZA,PABLO,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: FRANCO,GREGORY
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 01/20/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:52

SUBJECT'S LAST NAME: PARKS

DATE/TIME OF TEST: 01/29/2017 01:21:55

Diagnostics Check		OK	01:14
Air Blank		0.000	01:14
Control Test		0.078	01:15
Air Blank		0.000	01:15
Subject Sample	# 1	0.202	01:16
Breath Volume		2.027	01:16
Air Blank		0.000	01:16
Air Blank		0.000	01:18
Subject Sample	# 2	0.206	01:20
Breath Volume		1.246	01:20
Air Blank		0.000	01:20
Control Test		0.077	01:21
Air Blank		0.000	01:21
Diagnostics Check		OK	01:21

Data Download Date/Time: 02/14/2017 13:48

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESPINOZA,PABLO,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: VARGAS,DAEL
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 01/20/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:51

SUBJECT'S LAST NAME: VALENCIA

DATE/TIME OF TEST: 01/29/2017 01:39:24

Diagnostics Check		OK	01:32
Air Blank		0.000	01:33
Control Test		0.078	01:33
Air Blank		0.000	01:34
Subject Sample	# 1	0.226	01:34
Breath Volume		2.636	01:34
Air Blank		0.000	01:35
Air Blank		0.000	01:37
Subject Sample	# 2	0.214	01:37
Breath Volume		1.796	01:37
Air Blank		0.000	01:38
Control Test		0.077	01:38
Air Blank		0.000	01:39
Diagnostics Check		OK	01:39

Data Download Date/Time: 02/14/2017 13:48

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEY,LAZARO,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: ABREU,ESTEBAN
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 01/20/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:17

SUBJECT'S LAST NAME: VALDES

DATE/TIME OF TEST: 01/29/2017 02:50:58

Diagnostics Check		OK	02:44
Air Blank		0.000	02:44
Control Test		0.079	02:45
Air Blank		0.000	02:45
Subject Sample	# 1	0.096	02:46
Breath Volume		2.679	02:46
Air Blank		0.000	02:47
Air Blank		0.000	02:48
Subject Sample	# 2	0.094	02:49
Breath Volume		2.933	02:49
Air Blank		0.000	02:50
Control Test		0.078	02:50
Air Blank		0.000	02:50
Diagnostics Check		OK	02:50

Data Download Date/Time: 02/14/2017 13:48

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEY,LAZARO,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: VALDEZ,MICHAEL
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 01/20/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:43

SUBJECT'S LAST NAME: SOSA

DATE/TIME OF TEST: 01/29/2017 03:26:36

Diagnostics Check		OK	03:17
Air Blank		0.000	03:17
Control Test		0.077	03:17
Air Blank		0.000	03:18
Subject Sample	# 1	0.178	03:18
Breath Volume		2.218	03:18
Air Blank		0.000	03:19
Air Blank		0.000	03:21
Subject Sample	# 2	0.201	03:21
Breath Volume		2.855	03:21
Air Blank		0.000	03:22
Air Blank		0.000	03:24
Subject Sample	# 3	0.189	03:24
Breath Volume		2.125	03:24
Air Blank		0.000	03:25
Control Test		0.076	03:25
Air Blank		0.000	03:26
Diagnostics Check		OK	03:26

Data Download Date/Time: 02/14/2017 13:48

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZRAK,JOSEPH,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: AZRAK,JOSEPH
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 12/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: AVILES

DATE/TIME OF TEST: 01/04/2017 03:46:54

Diagnostics Check		OK	03:40
Air Blank		0.000	03:40
Control Test		0.078	03:40
Air Blank		0.000	03:41
Subject Sample	# 1	0.231	03:42
Breath Volume		2.238	03:42
Air Blank		0.000	03:43
Air Blank		0.000	03:44
Subject Sample	# 2	0.234	03:45
Breath Volume		2.453	03:45
Air Blank		0.000	03:45
Control Test		0.079	03:46
Air Blank		0.000	03:46
Diagnostics Check		OK	03:46

Data Download Date/Time: 01/06/2017 10:33

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZRAK,JOSEPH,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: AZRAK,JOSEPH
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 12/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 01/05/2017 02:57:26

Diagnostics Check		OK	02:53	
Air Blank		0.000	02:53	
Control Test		0.079	02:54	
Air Blank		0.000	02:54	
Subject Sample	# 1		02:56	Subject Test Refused
Breath Volume		0.000	02:56	
Air Blank		0.000	02:56	
Control Test		0.079	02:56	
Air Blank		0.000	02:57	
Diagnostics Check		OK	02:57	

Data Download Date/Time: 01/06/2017 10:33

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZRAK,JOSEPH,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: AZRAK,JOSEPH
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 01/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:28

SUBJECT'S LAST NAME: PIERRE

DATE/TIME OF TEST: 01/07/2017 05:07:35

Diagnostics Check		OK	05:01
Air Blank		0.000	05:01
Control Test		0.077	05:01
Air Blank		0.000	05:02
Subject Sample	# 1	0.172	05:03
Breath Volume		1.242	05:03
Air Blank		0.000	05:03
Air Blank		0.000	05:05
Subject Sample	# 2	0.184	05:06
Breath Volume		1.218	05:06
Air Blank		0.000	05:06
Control Test		0.077	05:06
Air Blank		0.000	05:07
Diagnostics Check		OK	05:07

Data Download Date/Time: 02/03/2017 09:12

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZRAK,JOSEPH,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: PACHECO,GEORGE
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 01/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:31

SUBJECT'S LAST NAME: CUPIDON

DATE/TIME OF TEST: 01/11/2017 03:13:59

Diagnostics Check		OK	03:05
Air Blank		0.000	03:06
Control Test		0.080	03:06
Air Blank		0.000	03:06
Subject Sample	# 1	0.139	03:09
Breath Volume		1.105	03:09
Air Blank		0.000	03:10
Air Blank		0.000	03:12
Subject Sample	# 2	0.145	03:12
Breath Volume		1.105	03:12
Air Blank		0.000	03:13
Control Test		0.078	03:13
Air Blank		0.000	03:13
Diagnostics Check		OK	03:13

Data Download Date/Time: 02/03/2017 09:12

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZRAK,JOSEPH,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: AZRAK,JOSEPH
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 01/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:06

SUBJECT'S LAST NAME: KING

DATE/TIME OF TEST: 01/13/2017 03:37:19

Diagnostics Check		OK	03:31
Air Blank		0.000	03:31
Control Test		0.079	03:31
Air Blank		0.000	03:32
Subject Sample	# 1	0.179	03:32
Breath Volume		1.402	03:32
Air Blank		0.000	03:33
Air Blank		0.000	03:35
Subject Sample	# 2	0.178	03:35
Breath Volume		1.828	03:35
Air Blank		0.000	03:36
Control Test		0.077	03:36
Air Blank		0.000	03:37
Diagnostics Check		OK	03:37

Data Download Date/Time: 02/03/2017 09:12

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZRAK,JOSEPH,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: AZRAK,JOSEPH
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 01/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 05:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:34

SUBJECT'S LAST NAME: PINTO

DATE/TIME OF TEST: 01/14/2017 06:05:39

Diagnostics Check	OK	06:04	
Air Blank	0.000	06:04	
Control Test	0.086	06:05	Control Outside Tolerance
Air Blank	0.000	06:05	

Data Download Date/Time: 02/03/2017 09:12

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZRAK,JOSEPH,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: AZRAK,JOSEPH
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 01/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 05:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:34

SUBJECT'S LAST NAME: PINTO

DATE/TIME OF TEST: 01/14/2017 06:17:13

Diagnostics Check		OK	06:07	
Air Blank		0.000	06:08	
Control Test		0.076	06:08	
Air Blank		0.000	06:08	
Subject Sample	# 1		06:11	No Sample Provided
Breath Volume		0.000	06:11	
Air Blank		0.000	06:12	
Air Blank		0.000	06:14	
Subject Sample	# 2		06:15	Subject Test Refused
Breath Volume		0.000	06:15	
Air Blank		0.000	06:16	
Control Test		0.079	06:16	
Air Blank		0.000	06:17	
Diagnostics Check		OK	06:17	

Data Download Date/Time: 02/03/2017 09:12

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZRAK,JOSEPH,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: AZRAK,JOSEPH
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 01/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: OBREGON

DATE/TIME OF TEST: 01/25/2017 03:41:34

Diagnostics Check		OK	03:34
Air Blank		0.000	03:34
Control Test		0.078	03:34
Air Blank		0.000	03:35
Subject Sample	# 1	0.175	03:36
Breath Volume		1.667	03:36
Air Blank		0.000	03:37
Air Blank		0.000	03:39
Subject Sample	# 2	0.186	03:39
Breath Volume		1.207	03:39
Air Blank		0.000	03:40
Control Test		0.078	03:40
Air Blank		0.000	03:41
Diagnostics Check		OK	03:41

Data Download Date/Time: 02/03/2017 09:12

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZRAK,JOSEPH,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: AZRAK,JOSEPH
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 01/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:04

SUBJECT'S LAST NAME: LOPEZ GARCIA

DATE/TIME OF TEST: 01/27/2017 01:40:19

Diagnostics Check		OK	01:36	
Air Blank		0.000	01:36	
Control Test		0.079	01:37	
Air Blank		0.000	01:37	
Subject Sample	# 1		01:38	Subject Test Refused
Breath Volume		0.000	01:38	
Air Blank		0.000	01:39	
Control Test		0.079	01:39	
Air Blank		0.000	01:40	
Diagnostics Check		OK	01:40	

Data Download Date/Time: 02/03/2017 09:12

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005810

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUACKENBUSH,BRET,
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: QUACKENBUSH,BRET
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 12/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:42

SUBJECT'S LAST NAME: SQUIRE

DATE/TIME OF TEST: 01/01/2017 03:09:07

Diagnostics Check		OK	03:02
Air Blank		0.000	03:02
Control Test		0.078	03:03
Air Blank		0.000	03:03
Subject Sample	# 1	0.087	03:04
Breath Volume		2.257	03:04
Air Blank		0.000	03:05
Air Blank		0.000	03:06
Subject Sample	# 2	0.083	03:07
Breath Volume		1.441	03:07
Air Blank		0.000	03:08
Control Test		0.078	03:08
Air Blank		0.000	03:09
Diagnostics Check		OK	03:09

Data Download Date/Time: 01/24/2017 10:15

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005810

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTLE,CORY,A
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: SUTTLE,CORY
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 12/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: SWITZER

DATE/TIME OF TEST: 01/01/2017 03:31:37

Diagnostics Check		OK	03:25
Air Blank		0.000	03:25
Control Test		0.078	03:25
Air Blank		0.000	03:26
Subject Sample	# 1	0.222	03:26
Breath Volume		2.769	03:26
Air Blank		0.000	03:27
Air Blank		0.000	03:29
Subject Sample	# 2	0.214	03:29
Breath Volume		2.226	03:29
Air Blank		0.000	03:30
Control Test		0.077	03:30
Air Blank		0.000	03:31
Diagnostics Check		OK	03:31

Data Download Date/Time: 01/24/2017 10:15

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005810

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KELLNER,MICHAEL,A
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: KELLNER,MICHAEL
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 12/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: MCCAULLEY

DATE/TIME OF TEST: 01/01/2017 03:44:10

Diagnostics Check		OK	03:37
Air Blank		0.000	03:37
Control Test		0.077	03:38
Air Blank		0.000	03:38
Subject Sample	# 1	0.200	03:39
Breath Volume		1.511	03:39
Air Blank		0.000	03:40
Air Blank		0.000	03:42
Subject Sample	# 2	0.189	03:42
Breath Volume		1.472	03:42
Air Blank		0.000	03:43
Control Test		0.077	03:43
Air Blank		0.000	03:44
Diagnostics Check		OK	03:44

Data Download Date/Time: 01/24/2017 10:15

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005810

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KELLNER,MICHAEL,A
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: KELLNER,MICHAEL
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 12/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 19:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:33

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 01/01/2017 21:12:06

Diagnostics Check		OK	21:03
Air Blank		0.000	21:04
Control Test		0.076	21:04
Air Blank		0.000	21:04
Subject Sample	# 1	0.121	21:06
Breath Volume		1.378	21:06
Air Blank		0.000	21:07
Air Blank		0.000	21:09
Subject Sample	# 2	0.126	21:10
Breath Volume		2.000	21:10
Air Blank		0.000	21:11
Control Test		0.076	21:11
Air Blank		0.000	21:12
Diagnostics Check		OK	21:12

Data Download Date/Time: 01/24/2017 10:15

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005810

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUACKENBUSH,BRET,C
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: QUACKENBUSH,BRET
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 12/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: ORMSBEE

DATE/TIME OF TEST: 01/02/2017 00:44:16

Diagnostics Check		OK	00:36
Air Blank		0.000	00:36
Control Test		0.077	00:36
Air Blank		0.000	00:37
Subject Sample	# 1	0.211	00:39
Breath Volume		1.726	00:39
Air Blank		0.000	00:39
Air Blank		0.000	00:41
Subject Sample	# 2	0.210	00:42
Breath Volume		1.671	00:42
Air Blank		0.000	00:43
Control Test		0.075	00:43
Air Blank		0.000	00:44
Diagnostics Check		OK	00:44

Data Download Date/Time: 01/24/2017 10:15

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005810

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUACKENBUSH,BRET,C
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: STAFFORD,BRIAN
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 12/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:07

SUBJECT'S LAST NAME: GRAHAM

DATE/TIME OF TEST: 01/02/2017 01:39:48

Diagnostics Check		OK	01:32
Air Blank		0.000	01:32
Control Test		0.077	01:33
Air Blank		0.000	01:33
Subject Sample	# 1	0.132	01:34
Breath Volume		1.593	01:34
Air Blank		0.000	01:35
Air Blank		0.000	01:37
Subject Sample	# 2	0.129	01:38
Breath Volume		1.757	01:38
Air Blank		0.000	01:38
Control Test		0.077	01:39
Air Blank		0.000	01:39
Diagnostics Check		OK	01:39

Data Download Date/Time: 01/24/2017 10:15

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005810

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTLE,CORY,A
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: HARRISON,NICK
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 12/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: PROB/PAROL
ARREST TIME: 18:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:40

SUBJECT'S LAST NAME: GARRETT

DATE/TIME OF TEST: 01/03/2017 19:16:08

Diagnostics Check		OK	19:02	
Air Blank		0.000	19:03	
Control Test		0.077	19:03	
Air Blank		0.000	19:04	
Subject Sample	# 1	0.265	19:06	
Breath Volume		1.191	19:06	
Air Blank		0.000	19:07	
Air Blank		0.000	19:08	
Subject Sample	# 2		19:11	No Sample Provided
Breath Volume		0.000	19:11	
Air Blank		0.000	19:12	
Air Blank		0.000	19:14	
Subject Sample	# 3		19:14	Subject Test Refused
Breath Volume		0.000	19:14	
Air Blank		0.000	19:15	
Control Test		0.077	19:15	
Air Blank		0.000	19:16	
Diagnostics Check		OK	19:16	

Data Download Date/Time: 01/24/2017 10:15

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005810

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUACKENBUSH,BRET,C
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: QUACKENBUSH,BRET
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 12/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: MCPHERSON

DATE/TIME OF TEST: 01/05/2017 03:01:21

Diagnostics Check		OK	02:49	
Air Blank		0.000	02:50	
Control Test		0.077	02:50	
Air Blank		0.000	02:51	
Subject Sample	# 1		02:54	No Sample Provided
Breath Volume		0.000	02:54	
Air Blank		0.000	02:54	
Air Blank		0.000	02:56	
Subject Sample	# 2		02:59	No Sample Provided
Breath Volume		0.000	02:59	
Air Blank		0.000	03:00	
Control Test		0.078	03:00	
Air Blank		0.000	03:01	
Diagnostics Check		OK	03:01	

Data Download Date/Time: 01/24/2017 10:15

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005810

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KELLNER,MICHAEL,A
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: KELLNER,MICHAEL
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 12/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: HICKEY

DATE/TIME OF TEST: 01/05/2017 04:55:24

Diagnostics Check		OK	04:43	
Air Blank		0.000	04:44	
Control Test		0.077	04:44	
Air Blank		0.000	04:45	
Subject Sample	# 1		04:48	No Sample Provided
Breath Volume		0.000	04:48	
Air Blank		0.000	04:48	
Air Blank		0.000	04:50	
Subject Sample	# 2		04:54	No Sample Provided
Breath Volume		0.000	04:54	
Air Blank		0.000	04:54	
Control Test		0.078	04:54	
Air Blank		0.000	04:55	
Diagnostics Check		OK	04:55	

Data Download Date/Time: 01/24/2017 10:15

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005810

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KELLNER,MICHAEL,A
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: KELLNER,MICHAEL
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 12/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:03

SUBJECT'S LAST NAME: CROSS

DATE/TIME OF TEST: 01/06/2017 04:02:05

Diagnostics Check		OK	03:55
Air Blank		0.000	03:55
Control Test		0.077	03:55
Air Blank		0.000	03:56
Subject Sample	# 1	0.107	03:57
Breath Volume		1.878	03:57
Air Blank		0.000	03:57
Air Blank		0.000	03:59
Subject Sample	# 2	0.115	04:00
Breath Volume		1.843	04:00
Air Blank		0.000	04:01
Control Test		0.076	04:01
Air Blank		0.000	04:02
Diagnostics Check		OK	04:02

Data Download Date/Time: 01/24/2017 10:15

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005810

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTLE,CORY,A
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: SUTTLE,CORY
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 12/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:25

SUBJECT'S LAST NAME: LOWTHER

DATE/TIME OF TEST: 01/06/2017 22:51:34

Diagnostics Check		OK	22:45
Air Blank		0.000	22:45
Control Test		0.078	22:45
Air Blank		0.000	22:46
Subject Sample	# 1	0.231	22:47
Breath Volume		2.562	22:47
Air Blank		0.000	22:47
Air Blank		0.000	22:49
Subject Sample	# 2	0.217	22:49
Breath Volume		2.679	22:49
Air Blank		0.000	22:50
Control Test		0.076	22:50
Air Blank		0.000	22:51
Diagnostics Check		OK	22:51

Data Download Date/Time: 01/24/2017 10:15

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005810

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTLE,CORY,A
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: SUTTLE,CORY
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 12/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: CURRY

DATE/TIME OF TEST: 01/07/2017 02:47:28

Diagnostics Check		OK	02:35	
Air Blank		0.000	02:36	
Control Test		0.077	02:36	
Air Blank		0.000	02:36	
Subject Sample	# 1	0.181	02:40	Volume Not Met
Breath Volume		0.128	02:40	
Air Blank		0.000	02:40	
Air Blank		0.000	02:42	
Subject Sample	# 2	0.198	02:42	
Breath Volume		1.859	02:42	
Air Blank		0.000	02:43	
Air Blank		0.000	02:45	
Subject Sample	# 3	0.215	02:45	
Breath Volume		1.632	02:45	
Air Blank		0.000	02:46	
Control Test		0.075	02:46	
Air Blank		0.000	02:47	
Diagnostics Check		OK	02:47	

Data Download Date/Time: 01/24/2017 10:15

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005810

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTLE,CORY,A
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: SUTTLE,CORY
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 12/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: REYNA

DATE/TIME OF TEST: 01/08/2017 03:12:50

Diagnostics Check		OK	03:06
Air Blank		0.000	03:06
Control Test		0.078	03:06
Air Blank		0.000	03:07
Subject Sample	# 1	0.102	03:08
Breath Volume		2.183	03:08
Air Blank		0.000	03:09
Air Blank		0.000	03:10
Subject Sample	# 2	0.097	03:11
Breath Volume		1.714	03:11
Air Blank		0.000	03:11
Control Test		0.079	03:12
Air Blank		0.000	03:12
Diagnostics Check		OK	03:12

Data Download Date/Time: 01/24/2017 10:15

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005810

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KELLNER,MICHAEL,A
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: KELLNER,MICHAEL
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 12/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: TERRY

DATE/TIME OF TEST: 01/10/2017 02:26:11

Diagnostics Check		OK	02:19
Air Blank		0.000	02:19
Control Test		0.078	02:20
Air Blank		0.000	02:20
Subject Sample	# 1	0.134	02:21
Breath Volume		2.082	02:21
Air Blank		0.000	02:22
Air Blank		0.000	02:24
Subject Sample	# 2	0.128	02:24
Breath Volume		1.863	02:24
Air Blank		0.000	02:25
Control Test		0.077	02:25
Air Blank		0.000	02:26
Diagnostics Check		OK	02:26

Data Download Date/Time: 01/24/2017 10:15

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005810

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MULDERRIG,SEAN,F
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: MULDERRIG,SEAN
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 12/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:19

SUBJECT'S LAST NAME: MC NEIL

DATE/TIME OF TEST: 01/11/2017 03:49:13

Diagnostics Check		OK	03:43
Air Blank		0.000	03:43
Control Test		0.078	03:43
Air Blank		0.000	03:44
Subject Sample	# 1	0.000	03:44
Breath Volume		1.160	03:44
Air Blank		0.000	03:45
Air Blank		0.000	03:47
Subject Sample	# 2	0.000	03:47
Breath Volume		1.785	03:47
Air Blank		0.000	03:48
Control Test		0.078	03:48
Air Blank		0.000	03:49
Diagnostics Check		OK	03:49

Data Download Date/Time: 01/24/2017 10:15

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005810

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTLE,CORY,A
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: SUTTLE,CORY
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 12/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: BUSTOS

DATE/TIME OF TEST: 01/12/2017 02:56:46

Diagnostics Check		OK	02:50
Air Blank		0.000	02:51
Control Test		0.077	02:51
Air Blank		0.000	02:51
Subject Sample	# 1	0.173	02:52
Breath Volume		2.316	02:52
Air Blank		0.000	02:52
Air Blank		0.000	02:54
Subject Sample	# 2	0.163	02:55
Breath Volume		1.941	02:55
Air Blank		0.000	02:55
Control Test		0.075	02:56
Air Blank		0.000	02:56
Diagnostics Check		OK	02:56

Data Download Date/Time: 01/24/2017 10:15

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005810

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MULDERRIG,SEAN,F
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: MULDERRIG,SEAN
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 12/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:38

SUBJECT'S LAST NAME: VARGAS

DATE/TIME OF TEST: 01/15/2017 03:02:47

Diagnostics Check	OK	03:02	
Air Blank		03:02	RFI Detect
Air Blank	0.000	03:02	

Data Download Date/Time: 01/24/2017 10:15

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005810

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MULDERRIG,SEAN,F
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: MULDERRIG,SEAN
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 12/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:38

SUBJECT'S LAST NAME: VARGAS

DATE/TIME OF TEST: 01/15/2017 03:12:12

Diagnostics Check		OK	03:05
Air Blank		0.000	03:05
Control Test		0.078	03:06
Air Blank		0.000	03:06
Subject Sample	# 1	0.147	03:07
Breath Volume		1.332	03:07
Air Blank		0.000	03:08
Air Blank		0.000	03:09
Subject Sample	# 2	0.143	03:10
Breath Volume		2.156	03:10
Air Blank		0.000	03:11
Control Test		0.077	03:11
Air Blank		0.000	03:12
Diagnostics Check		OK	03:12

Data Download Date/Time: 01/24/2017 10:15

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005810

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MULDERRIG,SEAN,F
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: MULDERRIG,SEAN
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 12/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: MCQUAIG

DATE/TIME OF TEST: 01/16/2017 00:03:37

Diagnostics Check	OK	00:02	
Air Blank	0.000	00:02	
Control Test	0.077	00:02	
Air Blank		00:03	RFI Detect
Air Blank	0.000	00:03	

Data Download Date/Time: 01/24/2017 10:15

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005810

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MULDERRIG,SEAN,F
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: MULDERRIG,SEAN
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 12/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: MCQUAIG

DATE/TIME OF TEST: 01/16/2017 00:16:29

Diagnostics Check		OK	00:05	
Air Blank		0.000	00:05	
Control Test		0.077	00:05	
Air Blank		0.000	00:06	
Subject Sample	# 1		00:09	No Sample Provided
Breath Volume		0.000	00:09	
Air Blank		0.000	00:09	
Air Blank		0.000	00:11	
Subject Sample	# 2		00:15	No Sample Provided
Breath Volume		0.000	00:15	
Air Blank		0.000	00:15	
Control Test		0.077	00:15	
Air Blank		0.000	00:16	
Diagnostics Check		OK	00:16	

Data Download Date/Time: 01/24/2017 10:15

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005810

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUACKENBUSH,BRET,C
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: QUACKENBUSH,BRET
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 12/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:06

SUBJECT'S LAST NAME: SIEDLECKI

DATE/TIME OF TEST: 01/16/2017 01:39:01

Diagnostics Check		OK	01:27	
Air Blank		0.000	01:28	
Control Test		0.077	01:28	
Air Blank		0.000	01:28	
Subject Sample	# 1		01:32	No Sample Provided
Breath Volume		0.000	01:32	
Air Blank		0.000	01:32	
Air Blank		0.000	01:34	
Subject Sample	# 2		01:37	No Sample Provided
Breath Volume		0.000	01:37	
Air Blank		0.000	01:38	
Control Test		0.078	01:38	
Air Blank		0.000	01:38	
Diagnostics Check		OK	01:39	

Data Download Date/Time: 01/24/2017 10:15

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005810

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUACKENBUSH,BRET,C
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: QUACKENBUSH,BRET
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 12/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 17:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:16

SUBJECT'S LAST NAME: BRYANT

DATE/TIME OF TEST: 01/18/2017 18:48:54

Diagnostics Check		OK	18:43	
Air Blank		0.000	18:44	
Control Test		0.078	18:44	
Air Blank		0.000	18:45	
Subject Sample	# 1		18:47	RFI Detect
Breath Volume		0.000	18:47	
Air Blank		0.000	18:47	
Control Test		0.078	18:48	
Air Blank		0.000	18:48	
Diagnostics Check		OK	18:48	

Data Download Date/Time: 01/24/2017 10:15

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005810

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUACKENBUSH,BRET,C
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: QUACKENBUSH,BRET
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 12/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 17:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:16

SUBJECT'S LAST NAME: BRYANT

DATE/TIME OF TEST: 01/18/2017 19:02:30

Diagnostics Check		OK	18:51	
Air Blank		0.000	18:51	
Control Test		0.078	18:51	
Air Blank		0.000	18:52	
Subject Sample	# 1		18:55	No Sample Provided
Breath Volume		0.000	18:55	
Air Blank		0.000	18:55	
Air Blank		0.000	18:57	
Subject Sample	# 2		19:01	No Sample Provided
Breath Volume		0.000	19:01	
Air Blank		0.000	19:01	
Control Test		0.079	19:01	
Air Blank		0.000	19:02	
Diagnostics Check		OK	19:02	

Data Download Date/Time: 01/24/2017 10:15

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005810

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MULDERRIG,SEAN,F
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: MULDERRIG,SEAN
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 12/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: WAWRZYK

DATE/TIME OF TEST: 01/20/2017 04:57:30

Diagnostics Check		OK	04:51
Air Blank		0.000	04:51
Control Test		0.078	04:51
Air Blank		0.000	04:52
Subject Sample	# 1	0.084	04:52
Breath Volume		1.660	04:52
Air Blank		0.000	04:53
Air Blank		0.000	04:55
Subject Sample	# 2	0.086	04:55
Breath Volume		2.343	04:55
Air Blank		0.000	04:56
Control Test		0.077	04:56
Air Blank		0.000	04:57
Diagnostics Check		OK	04:57

Data Download Date/Time: 01/24/2017 10:15

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005810

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTLE,CORY,A
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: SUTTLE,CORY
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 12/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: BUSTAMANTE

DATE/TIME OF TEST: 01/21/2017 03:43:03

Diagnostics Check		OK	03:36
Air Blank		0.000	03:36
Control Test		0.078	03:36
Air Blank		0.000	03:37
Subject Sample	# 1	0.205	03:37
Breath Volume		1.601	03:37
Air Blank		0.000	03:38
Air Blank		0.000	03:40
Subject Sample	# 2	0.214	03:41
Breath Volume		1.273	03:41
Air Blank		0.000	03:42
Control Test		0.076	03:42
Air Blank		0.000	03:42
Diagnostics Check		OK	03:43

Data Download Date/Time: 01/24/2017 10:15

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005810

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTLE,CORY,A
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: SUTTLE,CORY
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 12/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: ELLIOTT

DATE/TIME OF TEST: 01/22/2017 22:17:53

Diagnostics Check		OK	22:05	
Air Blank		0.000	22:05	
Control Test		0.078	22:06	
Air Blank		0.000	22:06	
Subject Sample	# 1	0.008	22:09	Volume Not Met
Breath Volume		0.492	22:09	
Air Blank		0.000	22:10	
Air Blank		0.000	22:12	
Subject Sample	# 2	0.010	22:12	
Breath Volume		1.179	22:12	
Air Blank		0.000	22:13	
Air Blank		0.000	22:15	
Subject Sample	# 3	0.008	22:16	
Breath Volume		1.296	22:16	
Air Blank		0.000	22:16	
Control Test		0.077	22:17	
Air Blank		0.000	22:17	
Diagnostics Check		OK	22:17	

Data Download Date/Time: 01/24/2017 10:15

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005810

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MULDERRIG,SEAN,F
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: MULDERRIG,SEAN
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 12/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:24

SUBJECT'S LAST NAME: BREWSTER

DATE/TIME OF TEST: 01/23/2017 06:05:11

Diagnostics Check		OK	05:58
Air Blank		0.000	05:58
Control Test		0.078	05:59
Air Blank		0.000	05:59
Subject Sample	# 1	0.134	06:00
Breath Volume		2.031	06:00
Air Blank		0.000	06:01
Air Blank		0.000	06:02
Subject Sample	# 2	0.130	06:03
Breath Volume		2.500	06:03
Air Blank		0.000	06:04
Control Test		0.077	06:04
Air Blank		0.000	06:05
Diagnostics Check		OK	06:05

Data Download Date/Time: 01/24/2017 10:15

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005810

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KELLNER,MICHAEL,A
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: KELLNER,MICHAEL
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 01/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: COTTIER

DATE/TIME OF TEST: 01/28/2017 01:27:11

Diagnostics Check		OK	01:20
Air Blank		0.000	01:20
Control Test		0.078	01:20
Air Blank		0.000	01:21
Subject Sample	# 1	0.084	01:22
Breath Volume		1.500	01:22
Air Blank		0.000	01:23
Air Blank		0.000	01:24
Subject Sample	# 2	0.086	01:25
Breath Volume		1.355	01:25
Air Blank		0.000	01:26
Control Test		0.077	01:26
Air Blank		0.000	01:27
Diagnostics Check		OK	01:27

Data Download Date/Time: 02/27/2017 09:31

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005810

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KELLNER,MICHAEL,A
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: KELLNER,MICHAEL
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 01/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: DE LEON

DATE/TIME OF TEST: 01/29/2017 04:13:00

Diagnostics Check		OK	04:01	
Air Blank		0.000	04:02	
Control Test		0.078	04:02	
Air Blank		0.000	04:02	
Subject Sample	# 1		04:06	No Sample Provided
Breath Volume		0.000	04:06	
Air Blank		0.000	04:06	
Air Blank		0.000	04:08	
Subject Sample	# 2		04:11	No Sample Provided
Breath Volume		0.000	04:11	
Air Blank		0.000	04:12	
Control Test		0.079	04:12	
Air Blank		0.000	04:12	
Diagnostics Check		OK	04:13	

Data Download Date/Time: 02/27/2017 09:31

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005810

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUACKENBUSH,BRET,C
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: QUACKENBUSH,BRET
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 01/24/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:41

SUBJECT'S LAST NAME: MORRISON

DATE/TIME OF TEST: 01/30/2017 03:13:28

Diagnostics Check		OK	03:02	
Air Blank		0.000	03:02	
Control Test		0.078	03:02	
Air Blank		0.000	03:03	
Subject Sample	# 1		03:06	No Sample Provided
Breath Volume		0.000	03:06	
Air Blank		0.000	03:06	
Air Blank		0.000	03:08	
Subject Sample	# 2		03:12	No Sample Provided
Breath Volume		0.000	03:12	
Air Blank		0.000	03:12	
Control Test		0.079	03:12	
Air Blank		0.000	03:13	
Diagnostics Check		OK	03:13	

Data Download Date/Time: 02/27/2017 09:31

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005936

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PUFAHL,G,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: PUFAHL,G
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 12/13/2016

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:49

SUBJECT'S LAST NAME: BOYCE

DATE/TIME OF TEST: 01/07/2017 00:16:31

Diagnostics Check		OK	00:09
Air Blank		0.000	00:10
Control Test		0.080	00:10
Air Blank		0.000	00:11
Subject Sample	# 1	0.128	00:11
Breath Volume		3.632	00:11
Air Blank		0.000	00:12
Air Blank		0.000	00:14
Subject Sample	# 2	0.125	00:14
Breath Volume		3.601	00:14
Air Blank		0.000	00:15
Control Test		0.078	00:15
Air Blank		0.000	00:16
Diagnostics Check		OK	00:16

Data Download Date/Time: 01/27/2017 09:23

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005936

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PUFAHL,G,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: PUFAHL,G
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 12/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: PAYNE

DATE/TIME OF TEST: 01/13/2017 03:41:59

Diagnostics Check		OK	03:35
Air Blank		0.000	03:35
Control Test		0.081	03:36
Air Blank		0.000	03:36
Subject Sample	# 1	0.215	03:37
Breath Volume		2.531	03:37
Air Blank		0.000	03:38
Air Blank		0.000	03:39
Subject Sample	# 2	0.209	03:40
Breath Volume		1.324	03:40
Air Blank		0.000	03:41
Control Test		0.079	03:41
Air Blank		0.000	03:41
Diagnostics Check		OK	03:41

Data Download Date/Time: 01/27/2017 09:23

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005936

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PUFAHL,G,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: PUFAHL,G
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 12/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:31

SUBJECT'S LAST NAME: WEHAGEN

DATE/TIME OF TEST: 01/19/2017 20:58:41

Diagnostics Check		OK	20:52
Air Blank		0.000	20:52
Control Test		0.080	20:52
Air Blank		0.000	20:53
Subject Sample	# 1	0.236	20:53
Breath Volume		3.593	20:53
Air Blank		0.000	20:54
Air Blank		0.000	20:56
Subject Sample	# 2	0.237	20:56
Breath Volume		3.640	20:56
Air Blank		0.000	20:57
Control Test		0.079	20:58
Air Blank		0.000	20:58
Diagnostics Check		OK	20:58

Data Download Date/Time: 01/27/2017 09:23

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005936

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PUFAHL,G,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: PUFAHL,G
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 12/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:54

SUBJECT'S LAST NAME: STARLING

DATE/TIME OF TEST: 01/24/2017 19:20:49

Diagnostics Check		OK	19:14
Air Blank		0.000	19:15
Control Test		0.081	19:15
Air Blank		0.000	19:15
Subject Sample	# 1	0.283	19:16
Breath Volume		2.710	19:16
Air Blank		0.000	19:16
Air Blank		0.000	19:18
Subject Sample	# 2	0.281	19:19
Breath Volume		2.683	19:19
Air Blank		0.000	19:19
Control Test		0.080	19:20
Air Blank		0.000	19:20
Diagnostics Check		OK	19:20

Data Download Date/Time: 01/27/2017 09:23

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005936

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PUFAHL,G,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: PUFAHL,G
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 12/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:06

SUBJECT'S LAST NAME: MONROE

DATE/TIME OF TEST: 01/26/2017 04:34:29

Diagnostics Check		OK	04:26
Air Blank		0.000	04:26
Control Test		0.081	04:27
Air Blank		0.000	04:27
Subject Sample	# 1	0.197	04:29
Breath Volume		1.941	04:29
Air Blank		0.000	04:30
Air Blank		0.000	04:32
Subject Sample	# 2	0.198	04:32
Breath Volume		1.730	04:32
Air Blank		0.000	04:33
Control Test		0.078	04:33
Air Blank		0.000	04:34
Diagnostics Check		OK	04:34

Data Download Date/Time: 01/27/2017 09:23

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005936

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLINT,SERGEANT,C
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: PUFAHL,DEPUTY
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 01/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:03

SUBJECT'S LAST NAME: SAMPSON

DATE/TIME OF TEST: 01/27/2017 20:30:57

Diagnostics Check		OK	20:23
Air Blank		0.000	20:24
Control Test		0.082	20:24
Air Blank		0.000	20:24
Subject Sample	# 1	0.222	20:26
Breath Volume		3.132	20:26
Air Blank		0.000	20:26
Air Blank		0.000	20:28
Subject Sample	# 2	0.218	20:29
Breath Volume		3.070	20:29
Air Blank		0.000	20:30
Control Test		0.081	20:30
Air Blank		0.000	20:30
Diagnostics Check		OK	20:30

Data Download Date/Time: 02/27/2017 08:23

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SERRAES,BRANDON
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 01/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:25

SUBJECT'S LAST NAME: GROVER

DATE/TIME OF TEST: 01/13/2017 22:16:12

Diagnostics Check		OK	22:09
Air Blank		0.000	22:09
Control Test		0.081	22:10
Air Blank		0.000	22:10
Subject Sample	# 1	0.206	22:11
Breath Volume		2.164	22:11
Air Blank		0.000	22:11
Air Blank		0.000	22:13
Subject Sample	# 2	0.200	22:14
Breath Volume		2.097	22:14
Air Blank		0.000	22:15
Control Test		0.081	22:15
Air Blank		0.000	22:16
Diagnostics Check		OK	22:16

Data Download Date/Time: 02/10/2017 14:03

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DORSEY,PAMELA,F
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: PANIAGUA,MIGUEL
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 01/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:17

SUBJECT'S LAST NAME: GIVENS

DATE/TIME OF TEST: 01/14/2017 00:55:13

Diagnostics Check		OK	00:47
Air Blank		0.000	00:47
Control Test		0.081	00:48
Air Blank		0.000	00:48
Subject Sample	# 1	0.174	00:49
Breath Volume		1.367	00:49
Air Blank		0.000	00:50
Air Blank		0.000	00:52
Subject Sample	# 2	0.177	00:53
Breath Volume		1.359	00:53
Air Blank		0.000	00:54
Control Test		0.080	00:54
Air Blank		0.000	00:55
Diagnostics Check		OK	00:55

Data Download Date/Time: 02/10/2017 14:03

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITE,JENNIFER
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 01/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:57

SUBJECT'S LAST NAME: FORTUNE

DATE/TIME OF TEST: 01/14/2017 06:33:55

Diagnostics Check		OK	06:27
Air Blank		0.000	06:27
Control Test		0.081	06:28
Air Blank		0.000	06:28
Subject Sample	# 1	0.167	06:29
Breath Volume		1.453	06:29
Air Blank		0.000	06:30
Air Blank		0.000	06:31
Subject Sample	# 2	0.166	06:32
Breath Volume		1.832	06:32
Air Blank		0.000	06:32
Control Test		0.080	06:33
Air Blank		0.000	06:33
Diagnostics Check		OK	06:33

Data Download Date/Time: 02/10/2017 14:03

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: HOWARD,EMMA
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 01/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 06:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:30

SUBJECT'S LAST NAME: SANCHEZ LOPEZ

DATE/TIME OF TEST: 01/14/2017 08:06:12

Diagnostics Check		OK	07:58
Air Blank		0.000	07:59
Control Test		0.080	07:59
Air Blank		0.000	08:00
Subject Sample	# 1	0.114	08:01
Breath Volume		1.996	08:01
Air Blank		0.000	08:02
Air Blank		0.000	08:03
Subject Sample	# 2	0.112	08:04
Breath Volume		1.546	08:04
Air Blank		0.000	08:05
Control Test		0.081	08:05
Air Blank		0.000	08:06
Diagnostics Check		OK	08:06

Data Download Date/Time: 02/10/2017 14:03

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DORSEY,PAMELA,F
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: STEPHAN,ROBERT
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 01/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:27

SUBJECT'S LAST NAME: WELENC

DATE/TIME OF TEST: 01/14/2017 22:08:10

Diagnostics Check		OK	22:00
Air Blank		0.000	22:01
Control Test		0.081	22:01
Air Blank		0.000	22:01
Subject Sample	# 1	0.370	22:03
Breath Volume		4.351	22:03
Air Blank		0.000	22:04
Air Blank		0.000	22:05
Subject Sample	# 2	0.375	22:06
Breath Volume		3.406	22:06
Air Blank		0.000	22:07
Control Test		0.079	22:07
Air Blank		0.000	22:08
Diagnostics Check		OK	22:08

Data Download Date/Time: 02/10/2017 14:03

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DORSEY,PAMELA,F
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LYNCH,GREGORY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 01/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:59

SUBJECT'S LAST NAME: MCVICKER

DATE/TIME OF TEST: 01/15/2017 02:39:44

Diagnostics Check		OK	02:32
Air Blank		0.000	02:32
Control Test		0.081	02:33
Air Blank		0.000	02:33
Subject Sample	# 1	0.179	02:34
Breath Volume		2.406	02:34
Air Blank		0.000	02:35
Air Blank		0.000	02:37
Subject Sample	# 2	0.179	02:38
Breath Volume		1.289	02:38
Air Blank		0.000	02:38
Control Test		0.081	02:39
Air Blank		0.000	02:39
Diagnostics Check		OK	02:39

Data Download Date/Time: 02/10/2017 14:03

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ISHAM,MATTHEW
ARRESTING OFFICER AGENCY: Juno Beach PD
LAST AGENCY INSPECTION: 01/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: PHILLIPS

DATE/TIME OF TEST: 01/16/2017 02:24:54

Diagnostics Check		OK	02:18
Air Blank		0.000	02:18
Control Test		0.081	02:19
Air Blank		0.000	02:19
Subject Sample	# 1	0.253	02:20
Breath Volume		3.308	02:20
Air Blank		0.000	02:20
Air Blank		0.000	02:22
Subject Sample	# 2	0.254	02:23
Breath Volume		3.664	02:23
Air Blank		0.000	02:23
Control Test		0.080	02:24
Air Blank		0.000	02:24
Diagnostics Check		OK	02:24

Data Download Date/Time: 02/10/2017 14:03

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: KELLY,MAX
ARRESTING OFFICER AGENCY: Palm Beach PD
LAST AGENCY INSPECTION: 01/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 05:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:35

SUBJECT'S LAST NAME: HYMAN

DATE/TIME OF TEST: 01/16/2017 06:05:29

Diagnostics Check		OK	05:59
Air Blank		0.000	05:59
Control Test		0.082	06:00
Air Blank		0.000	06:00
Subject Sample	# 1	0.000	06:01
Breath Volume		2.468	06:01
Air Blank		0.000	06:01
Air Blank		0.000	06:03
Subject Sample	# 2	0.000	06:04
Breath Volume		2.625	06:04
Air Blank		0.000	06:04
Control Test		0.081	06:04
Air Blank		0.000	06:05
Diagnostics Check		OK	06:05

Data Download Date/Time: 02/10/2017 14:03

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SCHNEIDER,JOSHUA
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 01/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: FEENY

DATE/TIME OF TEST: 01/17/2017 03:08:29

Diagnostics Check		OK	03:02
Air Blank		0.000	03:02
Control Test		0.082	03:03
Air Blank		0.000	03:03
Subject Sample	# 1	0.138	03:04
Breath Volume		1.863	03:04
Air Blank		0.000	03:04
Air Blank		0.000	03:06
Subject Sample	# 2	0.137	03:06
Breath Volume		1.582	03:06
Air Blank		0.000	03:07
Control Test		0.082	03:07
Air Blank		0.000	03:08
Diagnostics Check		OK	03:08

Data Download Date/Time: 02/10/2017 14:03

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BORROWS,ANDREW
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 01/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:35

SUBJECT'S LAST NAME: WILLHAUCK

DATE/TIME OF TEST: 01/17/2017 05:05:50

Diagnostics Check		OK	04:58
Air Blank		0.000	04:58
Control Test		0.082	04:59
Air Blank		0.000	04:59
Subject Sample	# 1	0.137	05:01
Breath Volume		1.308	05:01
Air Blank		0.000	05:01
Air Blank		0.000	05:03
Subject Sample	# 2	0.146	05:04
Breath Volume		1.652	05:04
Air Blank		0.000	05:04
Control Test		0.081	05:05
Air Blank		0.000	05:05
Diagnostics Check		OK	05:05

Data Download Date/Time: 02/10/2017 14:03

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DORSEY,PAMELA,F
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: DESANTIS,BRIAN
ARRESTING OFFICER AGENCY: Riviera Beach PD
LAST AGENCY INSPECTION: 01/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 14:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:50

SUBJECT'S LAST NAME: PELLMAN

DATE/TIME OF TEST: 01/17/2017 15:38:55

Diagnostics Check		OK	15:31
Air Blank		0.000	15:32
Control Test		0.082	15:32
Air Blank		0.000	15:32
Subject Sample	# 1	0.313	15:34
Breath Volume		2.375	15:34
Air Blank		0.000	15:34
Air Blank		0.000	15:36
Subject Sample	# 2	0.311	15:37
Breath Volume		2.449	15:37
Air Blank		0.000	15:37
Control Test		0.080	15:38
Air Blank		0.000	15:38
Diagnostics Check		OK	15:38

Data Download Date/Time: 02/10/2017 14:03

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DORSEY,PAMELA,F
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LEVEY,SCOTT
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 01/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 18:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:22

SUBJECT'S LAST NAME: BOCKER

DATE/TIME OF TEST: 01/17/2017 19:54:48

Diagnostics Check		OK	19:47
Air Blank		0.000	19:48
Control Test		0.081	19:48
Air Blank		0.000	19:48
Subject Sample	# 1	0.165	19:50
Breath Volume		1.875	19:50
Air Blank		0.000	19:50
Air Blank		0.000	19:52
Subject Sample	# 2	0.163	19:53
Breath Volume		1.871	19:53
Air Blank		0.000	19:53
Control Test		0.081	19:54
Air Blank		0.000	19:54
Diagnostics Check		OK	19:54

Data Download Date/Time: 02/10/2017 14:03

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: TODD,ZION
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 01/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: MONGATO

DATE/TIME OF TEST: 01/17/2017 22:25:56

Diagnostics Check		OK	22:19
Air Blank		0.000	22:19
Control Test		0.081	22:20
Air Blank		0.000	22:20
Subject Sample	# 1	0.000	22:21
Breath Volume		2.472	22:21
Air Blank		0.000	22:22
Air Blank		0.000	22:24
Subject Sample	# 2	0.000	22:24
Breath Volume		2.937	22:24
Air Blank		0.000	22:25
Control Test		0.082	22:25
Air Blank		0.000	22:25
Diagnostics Check		OK	22:25

Data Download Date/Time: 02/10/2017 14:03

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SCHAEFER,JOHN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 01/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:44

SUBJECT'S LAST NAME: KELLY

DATE/TIME OF TEST: 01/19/2017 03:15:12

Diagnostics Check		OK	03:07
Air Blank		0.000	03:08
Control Test		0.081	03:08
Air Blank		0.000	03:09
Subject Sample	# 1	0.171	03:10
Breath Volume		1.675	03:10
Air Blank		0.000	03:11
Air Blank		0.000	03:12
Subject Sample	# 2	0.178	03:13
Breath Volume		1.781	03:13
Air Blank		0.000	03:14
Control Test		0.080	03:14
Air Blank		0.000	03:15
Diagnostics Check		OK	03:15

Data Download Date/Time: 02/10/2017 14:03

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WARD,CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 01/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:17

SUBJECT'S LAST NAME: DELVA

DATE/TIME OF TEST: 01/21/2017 02:51:18

Diagnostics Check		OK	02:43
Air Blank		0.000	02:43
Control Test		0.082	02:43
Air Blank		0.000	02:44
Subject Sample	# 1	0.217	02:46
Breath Volume		1.777	02:46
Air Blank		0.000	02:46
Air Blank		0.000	02:48
Subject Sample	# 2	0.199	02:49
Breath Volume		1.382	02:49
Air Blank		0.000	02:50
Control Test		0.080	02:50
Air Blank		0.000	02:51
Diagnostics Check		OK	02:51

Data Download Date/Time: 02/10/2017 14:03

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARI,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: STEPHAN,ROBERT
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 01/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:57

SUBJECT'S LAST NAME: BARRERA JOVEL

DATE/TIME OF TEST: 01/22/2017 01:30:58

Diagnostics Check		OK	01:24
Air Blank		0.000	01:25
Control Test		0.081	01:25
Air Blank		0.000	01:25
Subject Sample	# 1	0.177	01:26
Breath Volume		2.167	01:26
Air Blank		0.000	01:27
Air Blank		0.000	01:28
Subject Sample	# 2	0.171	01:29
Breath Volume		3.363	01:29
Air Blank		0.000	01:30
Control Test		0.080	01:30
Air Blank		0.000	01:30
Diagnostics Check		OK	01:30

Data Download Date/Time: 02/10/2017 14:03

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARI,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WALTON,THOMAS
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 01/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: DELGADO JR

DATE/TIME OF TEST: 01/22/2017 02:49:49

Diagnostics Check		OK	02:42
Air Blank		0.000	02:42
Control Test		0.081	02:42
Air Blank		0.000	02:43
Subject Sample	# 1	0.093	02:45
Breath Volume		1.593	02:45
Air Blank		0.000	02:45
Air Blank		0.000	02:47
Subject Sample	# 2	0.091	02:48
Breath Volume		1.953	02:48
Air Blank		0.000	02:48
Control Test		0.080	02:49
Air Blank		0.000	02:49
Diagnostics Check		OK	02:49

Data Download Date/Time: 02/10/2017 14:03

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARI,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FRANKLIN,JONATHAN
ARRESTING OFFICER AGENCY: Tequesta PD
LAST AGENCY INSPECTION: 01/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:19

SUBJECT'S LAST NAME: LOWE

DATE/TIME OF TEST: 01/22/2017 04:56:42

Diagnostics Check		OK	04:49
Air Blank		0.000	04:49
Control Test		0.081	04:49
Air Blank		0.000	04:50
Subject Sample	# 1	0.209	04:51
Breath Volume		1.210	04:51
Air Blank		0.000	04:52
Air Blank		0.000	04:54
Subject Sample	# 2	0.212	04:55
Breath Volume		1.671	04:55
Air Blank		0.000	04:55
Control Test		0.079	04:56
Air Blank		0.000	04:56
Diagnostics Check		OK	04:56

Data Download Date/Time: 02/10/2017 14:03

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARI,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WEHRLE,JOHN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 01/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 11:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:50

SUBJECT'S LAST NAME: GOMEZ

DATE/TIME OF TEST: 01/22/2017 13:32:52

Diagnostics Check		OK	13:26
Air Blank		0.000	13:26
Control Test		0.081	13:27
Air Blank		0.000	13:27
Subject Sample	# 1	0.232	13:28
Breath Volume		1.375	13:28
Air Blank		0.000	13:29
Air Blank		0.000	13:30
Subject Sample	# 2	0.248	13:31
Breath Volume		1.500	13:31
Air Blank		0.000	13:31
Control Test		0.080	13:32
Air Blank		0.000	13:32
Diagnostics Check		OK	13:32

Data Download Date/Time: 02/10/2017 14:03

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITE,JENNIFER
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 01/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:55

SUBJECT'S LAST NAME: D'AMATO

DATE/TIME OF TEST: 01/23/2017 05:30:48

Diagnostics Check		OK	05:23
Air Blank		0.000	05:24
Control Test		0.082	05:24
Air Blank		0.000	05:25
Subject Sample	# 1	0.000	05:26
Breath Volume		4.394	05:26
Air Blank		0.000	05:26
Air Blank		0.000	05:28
Subject Sample	# 2	0.000	05:29
Breath Volume		3.734	05:29
Air Blank		0.000	05:29
Control Test		0.081	05:30
Air Blank		0.000	05:30
Diagnostics Check		OK	05:30

Data Download Date/Time: 02/10/2017 14:03

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DORSEY,PAMELA,F
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SCHNEIDER,DAVID
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 01/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:37

SUBJECT'S LAST NAME: CRUZ

DATE/TIME OF TEST: 01/23/2017 23:13:24

Diagnostics Check		OK	23:06
Air Blank		0.000	23:07
Control Test		0.081	23:07
Air Blank		0.000	23:08
Subject Sample	# 1	0.167	23:08
Breath Volume		2.316	23:08
Air Blank		0.000	23:09
Air Blank		0.000	23:11
Subject Sample	# 2	0.164	23:11
Breath Volume		2.898	23:11
Air Blank		0.000	23:12
Control Test		0.081	23:12
Air Blank		0.000	23:13
Diagnostics Check		OK	23:13

Data Download Date/Time: 02/10/2017 14:03

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARI,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SCHNEIDER,JOSHUA
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 01/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:49

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 01/24/2017 02:22:29

Diagnostics Check		OK	02:16
Air Blank		0.000	02:16
Control Test		0.081	02:17
Air Blank		0.000	02:17
Subject Sample	# 1	0.000	02:18
Breath Volume		1.691	02:18
Air Blank		0.000	02:18
Air Blank		0.000	02:20
Subject Sample	# 2	0.000	02:21
Breath Volume		1.210	02:21
Air Blank		0.000	02:21
Control Test		0.081	02:21
Air Blank		0.000	02:22
Diagnostics Check		OK	02:22

Data Download Date/Time: 02/10/2017 14:03

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DORSEY,PAMELA,F
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LYNCH,GREGORY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 01/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:56

SUBJECT'S LAST NAME: GARCIA-MORALES

DATE/TIME OF TEST: 01/24/2017 23:42:06

Diagnostics Check		OK	23:34
Air Blank		0.000	23:34
Control Test		0.081	23:35
Air Blank		0.000	23:35
Subject Sample	# 1	0.226	23:37
Breath Volume		1.816	23:37
Air Blank		0.000	23:38
Air Blank		0.000	23:39
Subject Sample	# 2	0.222	23:40
Breath Volume		2.164	23:40
Air Blank		0.000	23:41
Control Test		0.079	23:41
Air Blank		0.000	23:42
Diagnostics Check		OK	23:42

Data Download Date/Time: 02/10/2017 14:03

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARI,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: POINTU,PATRICK
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 01/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:35

SUBJECT'S LAST NAME: BODOR

DATE/TIME OF TEST: 01/25/2017 06:06:21

Diagnostics Check		OK	05:59
Air Blank		0.000	05:59
Control Test		0.081	06:00
Air Blank		0.000	06:00
Subject Sample	# 1	0.241	06:01
Breath Volume		1.785	06:01
Air Blank		0.000	06:02
Air Blank		0.000	06:03
Subject Sample	# 2	0.245	06:04
Breath Volume		1.609	06:04
Air Blank		0.000	06:05
Control Test		0.080	06:05
Air Blank		0.000	06:06
Diagnostics Check		OK	06:06

Data Download Date/Time: 02/10/2017 14:03

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DORSEY,PAMELA,F
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MARTINEZ,DEANNA
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 01/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 10:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:25

SUBJECT'S LAST NAME: CRUZ

DATE/TIME OF TEST: 01/26/2017 11:16:41

Diagnostics Check		OK	11:09
Air Blank		0.000	11:09
Control Test		0.082	11:10
Air Blank		0.000	11:10
Subject Sample	# 1	0.000	11:12
Breath Volume		1.410	11:12
Air Blank		0.000	11:12
Air Blank		0.000	11:14
Subject Sample	# 2	0.000	11:15
Breath Volume		1.605	11:15
Air Blank		0.000	11:15
Control Test		0.081	11:16
Air Blank		0.000	11:16
Diagnostics Check		OK	11:16

Data Download Date/Time: 02/10/2017 14:03

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FERNANDEZ,RYAN
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 01/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:06

SUBJECT'S LAST NAME: CARVAJAL

DATE/TIME OF TEST: 01/27/2017 02:33:07

Diagnostics Check		OK	02:30	
Air Blank		0.000	02:30	
Control Test		0.081	02:30	
Air Blank		0.000	02:31	
Subject Sample	# 1		02:31	Subject Test Refused
Breath Volume		0.000	02:31	
Air Blank		0.000	02:32	
Control Test		0.081	02:32	
Air Blank		0.000	02:33	
Diagnostics Check		OK	02:33	

Data Download Date/Time: 02/10/2017 14:04

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WARD,CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 01/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:05

SUBJECT'S LAST NAME: JENSEN

DATE/TIME OF TEST: 01/28/2017 05:35:53

Diagnostics Check		OK	05:28
Air Blank		0.000	05:29
Control Test		0.081	05:29
Air Blank		0.000	05:30
Subject Sample	# 1	0.217	05:31
Breath Volume		1.828	05:31
Air Blank		0.000	05:31
Air Blank		0.000	05:33
Subject Sample	# 2	0.217	05:34
Breath Volume		1.292	05:34
Air Blank		0.000	05:34
Control Test		0.080	05:35
Air Blank		0.000	05:35
Diagnostics Check		OK	05:35

Data Download Date/Time: 02/10/2017 14:04

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SENTMANAT,ADOLFO
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 01/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: CHUNNU

DATE/TIME OF TEST: 01/29/2017 00:50:53

Diagnostics Check		OK	00:44
Air Blank		0.000	00:45
Control Test		0.081	00:45
Air Blank		0.000	00:45
Subject Sample	# 1	0.214	00:46
Breath Volume		2.820	00:46
Air Blank		0.000	00:47
Air Blank		0.000	00:48
Subject Sample	# 2	0.223	00:49
Breath Volume		1.304	00:49
Air Blank		0.000	00:49
Control Test		0.080	00:50
Air Blank		0.000	00:50
Diagnostics Check		OK	00:50

Data Download Date/Time: 02/10/2017 14:04

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FERNANDEZ,RYAN
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 01/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: BOURQUIN

DATE/TIME OF TEST: 01/29/2017 03:14:58

Diagnostics Check		OK	03:08
Air Blank		0.000	03:08
Control Test		0.082	03:09
Air Blank		0.000	03:09
Subject Sample	# 1	0.224	03:10
Breath Volume		2.234	03:10
Air Blank		0.000	03:10
Air Blank		0.000	03:12
Subject Sample	# 2	0.219	03:13
Breath Volume		2.324	03:13
Air Blank		0.000	03:14
Control Test		0.081	03:14
Air Blank		0.000	03:14
Diagnostics Check		OK	03:14

Data Download Date/Time: 02/10/2017 14:04

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DORSEY,PAMELA,F
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FALCON,JESSICA
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 01/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 08:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:07

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 01/29/2017 09:20:29

Diagnostics Check		OK	09:13
Air Blank		0.000	09:14
Control Test		0.081	09:14
Air Blank		0.000	09:14
Subject Sample	# 1	0.101	09:15
Breath Volume		4.757	09:15
Air Blank		0.000	09:16
Air Blank		0.000	09:18
Subject Sample	# 2	0.102	09:18
Breath Volume		4.691	09:18
Air Blank		0.000	09:19
Control Test		0.081	09:19
Air Blank		0.000	09:20
Diagnostics Check		OK	09:20

Data Download Date/Time: 02/10/2017 14:04

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FICARRA,CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 01/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:58

SUBJECT'S LAST NAME: MARTINEZ RAMOS

DATE/TIME OF TEST: 01/30/2017 00:29:02

Diagnostics Check		OK	00:22
Air Blank		0.000	00:23
Control Test		0.082	00:23
Air Blank		0.000	00:23
Subject Sample	# 1	0.177	00:24
Breath Volume		3.074	00:24
Air Blank		0.000	00:25
Air Blank		0.000	00:26
Subject Sample	# 2	0.163	00:27
Breath Volume		1.957	00:27
Air Blank		0.000	00:28
Control Test		0.083	00:28
Air Blank		0.000	00:28
Diagnostics Check		OK	00:29

Data Download Date/Time: 02/10/2017 14:04

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SCHNEIDER,JOSHUA
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 01/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: MARC

DATE/TIME OF TEST: 01/30/2017 01:49:10

Diagnostics Check		OK	01:43
Air Blank		0.000	01:43
Control Test		0.081	01:44
Air Blank		0.000	01:44
Subject Sample	# 1	0.000	01:45
Breath Volume		1.953	01:45
Air Blank		0.000	01:45
Air Blank		0.000	01:47
Subject Sample	# 2	0.000	01:47
Breath Volume		2.003	01:47
Air Blank		0.000	01:48
Control Test		0.081	01:48
Air Blank		0.000	01:49
Diagnostics Check		OK	01:49

Data Download Date/Time: 02/10/2017 14:04

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DORSEY,PAMELA,F
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SEAY,ROBERT
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 01/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 15:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:28

SUBJECT'S LAST NAME: DILLON

DATE/TIME OF TEST: 01/30/2017 16:15:00

Diagnostics Check		OK	16:06
Air Blank		0.000	16:06
Control Test		0.082	16:06
Air Blank		0.000	16:07
Subject Sample	# 1	0.205	16:10
Breath Volume		1.300	16:10
Air Blank		0.000	16:10
Air Blank		0.000	16:12
Subject Sample	# 2	0.203	16:13
Breath Volume		1.445	16:13
Air Blank		0.000	16:14
Control Test		0.079	16:14
Air Blank		0.000	16:14
Diagnostics Check		OK	16:15

Data Download Date/Time: 02/10/2017 14:04

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DORSEY,PAMELA,F
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: TORRES,KERWIN
ARRESTING OFFICER AGENCY: Riviera Beach PD
LAST AGENCY INSPECTION: 01/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:35

SUBJECT'S LAST NAME: EATON

DATE/TIME OF TEST: 01/30/2017 21:10:32

Diagnostics Check		OK	21:03
Air Blank		0.000	21:03
Control Test		0.082	21:04
Air Blank		0.000	21:04
Subject Sample	# 1	0.173	21:05
Breath Volume		3.164	21:05
Air Blank		0.000	21:06
Air Blank		0.000	21:08
Subject Sample	# 2	0.162	21:08
Breath Volume		2.000	21:08
Air Blank		0.000	21:09
Control Test		0.081	21:09
Air Blank		0.000	21:10
Diagnostics Check		OK	21:10

Data Download Date/Time: 02/10/2017 14:04

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LOPEZ,BRANDON
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 01/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: RAMIREZ

DATE/TIME OF TEST: 01/31/2017 00:46:06

Diagnostics Check		OK	00:38
Air Blank		0.000	00:39
Control Test		0.082	00:39
Air Blank		0.000	00:40
Subject Sample	# 1	0.234	00:41
Breath Volume		3.078	00:41
Air Blank		0.000	00:42
Air Blank		0.000	00:43
Subject Sample	# 2	0.242	00:44
Breath Volume		2.929	00:44
Air Blank		0.000	00:45
Control Test		0.081	00:45
Air Blank		0.000	00:46
Diagnostics Check		OK	00:46

Data Download Date/Time: 02/10/2017 14:04

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DORSEY,PAMELA,F
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CHIQUITO-RODRIGUEZ,LENNDY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 01/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:53
GENDER: Female

OBSERVATION PERIOD BEGAN: 21:42

SUBJECT'S LAST NAME: PAPPAS

DATE/TIME OF TEST: 01/31/2017 22:14:09

Diagnostics Check		OK	22:07
Air Blank		0.000	22:07
Control Test		0.082	22:07
Air Blank		0.000	22:08
Subject Sample	# 1	0.000	22:09
Breath Volume		1.191	22:09
Air Blank		0.000	22:10
Air Blank		0.000	22:12
Subject Sample	# 2	0.000	22:12
Breath Volume		1.242	22:12
Air Blank		0.000	22:13
Control Test		0.081	22:13
Air Blank		0.000	22:14
Diagnostics Check		OK	22:14

Data Download Date/Time: 02/10/2017 14:04

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARI,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: POINTU,PATRICK
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 01/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:21

SUBJECT'S LAST NAME: DEDOMINICO

DATE/TIME OF TEST: 01/31/2017 23:55:49

Diagnostics Check		OK	23:48
Air Blank		0.000	23:49
Control Test		0.081	23:49
Air Blank		0.000	23:50
Subject Sample	# 1	0.157	23:51
Breath Volume		1.285	23:51
Air Blank		0.000	23:52
Air Blank		0.000	23:53
Subject Sample	# 2	0.155	23:54
Breath Volume		1.742	23:54
Air Blank		0.000	23:54
Control Test		0.078	23:55
Air Blank		0.000	23:55
Diagnostics Check		OK	23:55

Data Download Date/Time: 02/10/2017 14:04

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006046

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,ALAIN,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: HERNANDEZ,ALAIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 23:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:49

SUBJECT'S LAST NAME: SILVA

DATE/TIME OF TEST: 01/28/2017 00:15:55

Diagnostics Check		OK	00:09
Air Blank		0.000	00:10
Control Test		0.080	00:10
Air Blank		0.000	00:10
Subject Sample	# 1	0.143	00:11
Breath Volume		1.535	00:11
Air Blank		0.000	00:12
Air Blank		0.000	00:14
Subject Sample	# 2	0.144	00:14
Breath Volume		1.382	00:14
Air Blank		0.000	00:14
Control Test		0.081	00:15
Air Blank		0.000	00:15
Diagnostics Check		OK	00:15

Data Download Date/Time: 02/28/2017 00:00

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006046

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOKENDORF,BRIAN,K
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: DOKENDORF,BRIAN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 19:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:40

SUBJECT'S LAST NAME: BELLNIER

DATE/TIME OF TEST: 01/28/2017 20:36:21

Diagnostics Check		OK	20:30
Air Blank		0.000	20:30
Control Test		0.081	20:30
Air Blank		0.000	20:31
Subject Sample	# 1	0.109	20:31
Breath Volume		1.277	20:31
Air Blank		0.000	20:32
Air Blank		0.000	20:34
Subject Sample	# 2	0.090	20:34
Breath Volume		1.140	20:34
Air Blank		0.000	20:35
Control Test		0.079	20:35
Air Blank		0.000	20:36
Diagnostics Check		OK	20:36

Data Download Date/Time: 02/28/2017 00:00

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006168

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 12/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A2
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: O'MEARA

DATE/TIME OF TEST: 01/01/2017 02:36:59

Diagnostics Check		OK	02:28	
Air Blank		0.000	02:29	
Control Test		0.079	02:29	
Air Blank		0.000	02:29	
Subject Sample	# 1	0.194	02:32	Volume Not Met
Breath Volume		0.000	02:32	
Air Blank		0.000	02:33	
Air Blank		0.000	02:35	
Subject Sample	# 2		02:35	Subject Test Refused
Breath Volume		0.000	02:35	
Air Blank		0.000	02:36	
Control Test		0.077	02:36	
Air Blank		0.000	02:36	
Diagnostics Check		OK	02:36	

Data Download Date/Time: 01/20/2017 05:46

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006168

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEE,COREY,C
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: LEE,C
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 12/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A2
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: MCDONALD

DATE/TIME OF TEST: 01/01/2017 03:32:27

Diagnostics Check		OK	03:26
Air Blank		0.000	03:26
Control Test		0.078	03:26
Air Blank		0.000	03:27
Subject Sample	# 1	0.181	03:27
Breath Volume		1.277	03:27
Air Blank		0.000	03:28
Air Blank		0.000	03:30
Subject Sample	# 2	0.183	03:30
Breath Volume		1.976	03:30
Air Blank		0.000	03:31
Control Test		0.077	03:31
Air Blank		0.000	03:32
Diagnostics Check		OK	03:32

Data Download Date/Time: 01/20/2017 05:46

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006168

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PECCI,Z,A
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: PECCI,Z
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 12/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A2
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 01/01/2017 05:11:43

Diagnostics Check		OK	05:05
Air Blank		0.000	05:05
Control Test		0.079	05:05
Air Blank		0.000	05:06
Subject Sample	# 1	0.112	05:06
Breath Volume		3.781	05:06
Air Blank		0.000	05:07
Air Blank		0.000	05:09
Subject Sample	# 2	0.109	05:10
Breath Volume		3.664	05:10
Air Blank		0.000	05:10
Control Test		0.079	05:11
Air Blank		0.000	05:11
Diagnostics Check		OK	05:11

Data Download Date/Time: 01/20/2017 05:46

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006168

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CUTY,J,K
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: CUTY,J
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 12/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A2
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 12:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:57

SUBJECT'S LAST NAME: RAMIREZ-FUNES

DATE/TIME OF TEST: 01/01/2017 13:26:11

Diagnostics Check		OK	13:17
Air Blank		0.000	13:17
Control Test		0.080	13:18
Air Blank		0.000	13:18
Subject Sample	# 1	0.305	13:19
Breath Volume		2.343	13:19
Air Blank		0.000	13:19
Air Blank		0.000	13:21
Subject Sample	# 2	0.256	13:21
Breath Volume		1.277	13:21
Air Blank		0.000	13:22
Air Blank		0.000	13:24
Subject Sample	# 3	0.262	13:24
Breath Volume		1.394	13:24
Air Blank		0.000	13:25
Control Test		0.079	13:25
Air Blank		0.000	13:26
Diagnostics Check		OK	13:26

Data Download Date/Time: 01/20/2017 05:46

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006168

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDONALD,THOMAS,C
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: MCDONALD,THOMAS
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 12/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A2
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:22

SUBJECT'S LAST NAME: EDWARDS

DATE/TIME OF TEST: 01/03/2017 00:48:53

Diagnostics Check		OK	00:42
Air Blank		0.000	00:43
Control Test		0.079	00:43
Air Blank		0.000	00:43
Subject Sample	# 1	0.335	00:44
Breath Volume		2.773	00:44
Air Blank		0.000	00:45
Air Blank		0.000	00:46
Subject Sample	# 2	0.327	00:47
Breath Volume		1.277	00:47
Air Blank		0.000	00:47
Control Test		0.078	00:48
Air Blank		0.000	00:48
Diagnostics Check		OK	00:48

Data Download Date/Time: 01/20/2017 05:47

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006168

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CUTY,J,K
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: CUTY,J
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 12/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A2
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 13:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:54

SUBJECT'S LAST NAME: NICLAS

DATE/TIME OF TEST: 01/04/2017 14:21:55

Diagnostics Check		OK	14:15
Air Blank		0.000	14:16
Control Test		0.080	14:16
Air Blank		0.000	14:17
Subject Sample	# 1	0.286	14:17
Breath Volume		1.933	14:17
Air Blank		0.000	14:18
Air Blank		0.000	14:20
Subject Sample	# 2	0.279	14:20
Breath Volume		2.062	14:20
Air Blank		0.000	14:21
Control Test		0.079	14:21
Air Blank		0.000	14:21
Diagnostics Check		OK	14:21

Data Download Date/Time: 01/20/2017 05:47

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006168

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELLAS,DILLON,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: ROBINSON,ROBERT
ARRESTING OFFICER AGENCY: Tequesta PD
LAST AGENCY INSPECTION: 12/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A2
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:42

SUBJECT'S LAST NAME: JAMES

DATE/TIME OF TEST: 01/07/2017 04:24:46

Diagnostics Check		OK	04:18
Air Blank		0.000	04:19
Control Test		0.079	04:19
Air Blank		0.000	04:19
Subject Sample	# 1	0.175	04:20
Breath Volume		2.441	04:20
Air Blank		0.000	04:20
Air Blank		0.000	04:22
Subject Sample	# 2	0.172	04:23
Breath Volume		2.269	04:23
Air Blank		0.000	04:23
Control Test		0.078	04:24
Air Blank		0.000	04:24
Diagnostics Check		OK	04:24

Data Download Date/Time: 01/20/2017 05:47

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006168

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POLIZZI,CHRISTINE,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: POLIZZI,CHRISTINE
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 12/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A2
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 18:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:20

SUBJECT'S LAST NAME: SCOTT

DATE/TIME OF TEST: 01/07/2017 18:47:14

Diagnostics Check		OK	18:40
Air Blank		0.000	18:41
Control Test		0.079	18:41
Air Blank		0.000	18:42
Subject Sample	# 1	0.232	18:42
Breath Volume		2.585	18:42
Air Blank		0.000	18:43
Air Blank		0.000	18:45
Subject Sample	# 2	0.228	18:45
Breath Volume		2.121	18:45
Air Blank		0.000	18:46
Control Test		0.079	18:46
Air Blank		0.000	18:47
Diagnostics Check		OK	18:47

Data Download Date/Time: 01/20/2017 05:47

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006168

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELLAS,DILLON,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: BELLAS,DILLON
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 12/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A2
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 19:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:36

SUBJECT'S LAST NAME: BATTE

DATE/TIME OF TEST: 01/07/2017 20:03:12

Diagnostics Check		OK	19:56
Air Blank		0.000	19:57
Control Test		0.079	19:57
Air Blank		0.000	19:58
Subject Sample	# 1	0.354	19:58
Breath Volume		1.941	19:58
Air Blank		0.000	19:59
Air Blank		0.000	20:01
Subject Sample	# 2	0.363	20:01
Breath Volume		2.105	20:01
Air Blank		0.000	20:02
Control Test		0.080	20:02
Air Blank		0.000	20:03
Diagnostics Check		OK	20:03

Data Download Date/Time: 01/20/2017 05:47

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006168

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUEDIGER,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: RUEDIGER,CHRISTOPHER
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 12/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A2
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:54

SUBJECT'S LAST NAME: MARAGH

DATE/TIME OF TEST: 01/08/2017 02:09:16

Diagnostics Check		OK	01:54	
Air Blank		0.000	01:54	
Control Test		0.079	01:54	
Air Blank		0.000	01:55	
Subject Sample	# 1	0.177	01:58	
Breath Volume		1.750	01:58	
Air Blank		0.000	01:59	
Air Blank		0.000	02:01	
Subject Sample	# 2	0.174	02:04	Volume Not Met
Breath Volume		0.535	02:04	
Air Blank		0.000	02:05	
Air Blank		0.000	02:06	
Subject Sample	# 3	0.177	02:07	
Breath Volume		2.097	02:07	
Air Blank		0.000	02:08	
Control Test		0.077	02:08	
Air Blank		0.000	02:09	
Diagnostics Check		OK	02:09	

Data Download Date/Time: 01/20/2017 05:47

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006168

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAY,ERICK,J
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: DAY,ERICK
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 12/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A2
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:12

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 01/10/2017 03:01:44

Diagnostics Check		OK	02:54
Air Blank		0.000	02:54
Control Test		0.080	02:54
Air Blank		0.000	02:55
Subject Sample	# 1	0.168	02:56
Breath Volume		1.550	02:56
Air Blank		0.000	02:56
Air Blank		0.000	02:58
Subject Sample	# 2	0.171	03:00
Breath Volume		1.574	03:00
Air Blank		0.000	03:00
Control Test		0.079	03:01
Air Blank		0.000	03:01
Diagnostics Check		OK	03:01

Data Download Date/Time: 01/20/2017 05:47

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006168

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAY,ERICK,J
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: DAY,ERICK
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 12/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A2
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:32

SUBJECT'S LAST NAME: RAO

DATE/TIME OF TEST: 01/10/2017 21:41:38

Diagnostics Check		OK	21:33
Air Blank		0.000	21:33
Control Test		0.079	21:34
Air Blank		0.000	21:34
Subject Sample	# 1	0.153	21:36
Breath Volume		1.535	21:36
Air Blank		0.000	21:36
Air Blank		0.000	21:38
Subject Sample	# 2	0.164	21:40
Breath Volume		1.195	21:40
Air Blank		0.000	21:40
Control Test		0.077	21:41
Air Blank		0.000	21:41
Diagnostics Check		OK	21:41

Data Download Date/Time: 01/20/2017 05:47

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006168

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAKER,BRADFORD,B
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: BAKER,BRADFORD
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 12/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A2
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: ANDERSEN

DATE/TIME OF TEST: 01/14/2017 01:24:55

Diagnostics Check		OK	01:18
Air Blank		0.000	01:18
Control Test		0.079	01:19
Air Blank		0.000	01:19
Subject Sample	# 1	0.252	01:20
Breath Volume		1.859	01:20
Air Blank		0.000	01:20
Air Blank		0.000	01:22
Subject Sample	# 2	0.248	01:23
Breath Volume		2.207	01:23
Air Blank		0.000	01:24
Control Test		0.079	01:24
Air Blank		0.000	01:24
Diagnostics Check		OK	01:24

Data Download Date/Time: 01/20/2017 05:47

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006168

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VO,NGOAN,H
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: VO,NGOAN
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 12/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A2
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:18

SUBJECT'S LAST NAME: EHLERS

DATE/TIME OF TEST: 01/14/2017 20:53:19

Diagnostics Check		OK	20:47
Air Blank		0.000	20:47
Control Test		0.079	20:47
Air Blank		0.000	20:48
Subject Sample	# 1	0.301	20:48
Breath Volume		1.632	20:48
Air Blank		0.000	20:49
Air Blank		0.000	20:51
Subject Sample	# 2	0.299	20:51
Breath Volume		1.835	20:51
Air Blank		0.000	20:52
Control Test		0.078	20:52
Air Blank		0.000	20:53
Diagnostics Check		OK	20:53

Data Download Date/Time: 01/20/2017 05:47

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006168

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEELE,MARK,T
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: STEELE,MARK
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 12/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A2
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 15:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:15

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 01/15/2017 17:48:42

Diagnostics Check		OK	17:42
Air Blank		0.000	17:42
Control Test		0.082	17:43
Air Blank		0.000	17:43
Subject Sample	# 1	0.099	17:44
Breath Volume		3.375	17:44
Air Blank		0.000	17:44
Air Blank		0.000	17:46
Subject Sample	# 2	0.098	17:47
Breath Volume		3.605	17:47
Air Blank		0.000	17:47
Control Test		0.080	17:48
Air Blank		0.000	17:48
Diagnostics Check		OK	17:48

Data Download Date/Time: 01/20/2017 05:47

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006168

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POLIZZI,CHRISTINE,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: POLIZZI,CHRISTINE
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 12/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A2
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:23

SUBJECT'S LAST NAME: SUCHY

DATE/TIME OF TEST: 01/17/2017 23:51:22

Diagnostics Check		OK	23:44
Air Blank		0.000	23:45
Control Test		0.080	23:45
Air Blank		0.000	23:46
Subject Sample	# 1	0.126	23:46
Breath Volume		1.683	23:46
Air Blank		0.000	23:47
Air Blank		0.000	23:49
Subject Sample	# 2	0.119	23:49
Breath Volume		1.527	23:49
Air Blank		0.000	23:50
Control Test		0.080	23:50
Air Blank		0.000	23:51
Diagnostics Check		OK	23:51

Data Download Date/Time: 01/20/2017 05:47

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006168

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAKER,BRADFORD,B
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: BAKER,BRADFORD
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 12/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A2
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:47

SUBJECT'S LAST NAME: FRANCISCO

DATE/TIME OF TEST: 01/20/2017 00:20:37

Diagnostics Check		OK	00:14
Air Blank		0.000	00:14
Control Test		0.080	00:14
Air Blank		0.000	00:15
Subject Sample	# 1	0.227	00:15
Breath Volume		2.121	00:15
Air Blank		0.000	00:16
Air Blank		0.000	00:18
Subject Sample	# 2	0.231	00:19
Breath Volume		1.777	00:19
Air Blank		0.000	00:19
Control Test		0.079	00:20
Air Blank		0.000	00:20
Diagnostics Check		OK	00:20

Data Download Date/Time: 01/20/2017 05:47

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006168

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 01/20/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:56

SUBJECT'S LAST NAME: SOLIS-AILON

DATE/TIME OF TEST: 01/21/2017 01:23:25

Diagnostics Check		OK	01:16
Air Blank		0.000	01:17
Control Test		0.081	01:17
Air Blank		0.000	01:18
Subject Sample	# 1	0.177	01:18
Breath Volume		1.750	01:18
Air Blank		0.000	01:19
Air Blank		0.000	01:21
Subject Sample	# 2	0.169	01:21
Breath Volume		2.421	01:21
Air Blank		0.000	01:22
Control Test		0.080	01:22
Air Blank		0.000	01:23
Diagnostics Check		OK	01:23

Data Download Date/Time: 02/27/2017 13:36

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006168

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELLAS,DILLON,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: BELLAS,DILLON
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 01/20/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:08

SUBJECT'S LAST NAME: PAULY

DATE/TIME OF TEST: 01/21/2017 01:37:45

Diagnostics Check		OK	01:31
Air Blank		0.000	01:31
Control Test		0.080	01:31
Air Blank		0.000	01:32
Subject Sample	# 1	0.046	01:33
Breath Volume		3.746	01:33
Air Blank		0.000	01:33
Air Blank		0.000	01:35
Subject Sample	# 2	0.046	01:36
Breath Volume		3.949	01:36
Air Blank		0.000	01:36
Control Test		0.080	01:37
Air Blank		0.000	01:37
Diagnostics Check		OK	01:37

Data Download Date/Time: 02/27/2017 13:36

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006168

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 01/20/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:09

SUBJECT'S LAST NAME: AMAYA-SIFUENTES

DATE/TIME OF TEST: 01/22/2017 02:36:50

Diagnostics Check		OK	02:30
Air Blank		0.000	02:30
Control Test		0.081	02:31
Air Blank		0.000	02:31
Subject Sample	# 1	0.218	02:32
Breath Volume		2.785	02:32
Air Blank		0.000	02:32
Air Blank		0.000	02:34
Subject Sample	# 2	0.228	02:35
Breath Volume		3.667	02:35
Air Blank		0.000	02:35
Control Test		0.080	02:36
Air Blank		0.000	02:36
Diagnostics Check		OK	02:36

Data Download Date/Time: 02/27/2017 13:36

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006168

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEEDHAM,K,
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: LEEDHAM,K
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 01/20/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 18:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:55

SUBJECT'S LAST NAME: KOROUS

DATE/TIME OF TEST: 01/23/2017 19:24:05

Diagnostics Check		OK	19:17
Air Blank		0.000	19:18
Control Test		0.081	19:18
Air Blank		0.000	19:19
Subject Sample	# 1	0.355	19:19
Breath Volume		2.546	19:19
Air Blank		0.000	19:20
Air Blank		0.000	19:21
Subject Sample	# 2	0.358	19:22
Breath Volume		2.832	19:22
Air Blank		0.000	19:23
Control Test		0.081	19:23
Air Blank		0.000	19:24
Diagnostics Check		OK	19:24

Data Download Date/Time: 02/27/2017 13:36

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006168

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAY,ERICK,J
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: DAY,ERICK
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 01/20/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:12

SUBJECT'S LAST NAME: COSENTINO

DATE/TIME OF TEST: 01/24/2017 01:09:37

Diagnostics Check		OK	00:59
Air Blank		0.000	01:00
Control Test		0.080	01:00
Air Blank		0.000	01:01
Subject Sample	# 1	0.257	01:01
Breath Volume		3.367	01:01
Air Blank		0.000	01:02
Air Blank		0.000	01:04
Subject Sample	# 2	0.231	01:05
Breath Volume		3.085	01:05
Air Blank		0.000	01:05
Air Blank		0.000	01:07
Subject Sample	# 3	0.226	01:08
Breath Volume		2.980	01:08
Air Blank		0.000	01:08
Control Test		0.079	01:09
Air Blank		0.000	01:09
Diagnostics Check		OK	01:09

Data Download Date/Time: 02/27/2017 13:36

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006168

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAKER,BRADFORD,B
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: BAKER,BRADFORD
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 01/20/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 17:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:02

SUBJECT'S LAST NAME: MAYO

DATE/TIME OF TEST: 01/25/2017 18:33:12

Diagnostics Check		OK	18:27
Air Blank		0.000	18:27
Control Test		0.082	18:27
Air Blank		0.000	18:28
Subject Sample	# 1	0.039	18:28
Breath Volume		2.652	18:28
Air Blank		0.000	18:29
Air Blank		0.000	18:31
Subject Sample	# 2	0.038	18:31
Breath Volume		1.605	18:31
Air Blank		0.000	18:32
Control Test		0.081	18:32
Air Blank		0.000	18:33
Diagnostics Check		OK	18:33

Data Download Date/Time: 02/27/2017 13:36

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006168

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PECCI,Z,A
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: PECCI,Z
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 01/20/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 01/26/2017 02:57:23

Diagnostics Check		OK	02:50
Air Blank		0.000	02:51
Control Test		0.081	02:51
Air Blank		0.000	02:51
Subject Sample	# 1	0.154	02:52
Breath Volume		2.164	02:52
Air Blank		0.000	02:53
Air Blank		0.000	02:55
Subject Sample	# 2	0.153	02:55
Breath Volume		3.886	02:55
Air Blank		0.000	02:56
Control Test		0.080	02:56
Air Blank		0.000	02:57
Diagnostics Check		OK	02:57

Data Download Date/Time: 02/27/2017 13:36

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006168

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAKER,BRADFORD,B
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: BAKER,BRADFORD
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 01/20/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:11

SUBJECT'S LAST NAME: MAGNANO

DATE/TIME OF TEST: 01/27/2017 22:38:48

Diagnostics Check		OK	22:32
Air Blank		0.000	22:32
Control Test		0.081	22:33
Air Blank		0.000	22:33
Subject Sample	# 1	0.140	22:34
Breath Volume		2.664	22:34
Air Blank		0.000	22:34
Air Blank		0.000	22:36
Subject Sample	# 2	0.143	22:37
Breath Volume		2.648	22:37
Air Blank		0.000	22:37
Control Test		0.081	22:38
Air Blank		0.000	22:38
Diagnostics Check		OK	22:38

Data Download Date/Time: 02/27/2017 13:36

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006168

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANGERVIL,CLAUDY,P
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: ANGERVIL,CLAUDY
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 01/20/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:46

SUBJECT'S LAST NAME: LANSING

DATE/TIME OF TEST: 01/28/2017 01:13:25

Diagnostics Check		OK	01:07
Air Blank		0.000	01:07
Control Test		0.082	01:07
Air Blank		0.000	01:08
Subject Sample	# 1	0.095	01:08
Breath Volume		2.023	01:08
Air Blank		0.000	01:09
Air Blank		0.000	01:11
Subject Sample	# 2	0.094	01:11
Breath Volume		2.984	01:11
Air Blank		0.000	01:12
Control Test		0.080	01:12
Air Blank		0.000	01:13
Diagnostics Check		OK	01:13

Data Download Date/Time: 02/27/2017 13:36

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006168

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROOT,KATHRYN,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: ROOT,KATHRYN
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 01/20/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 14:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:18

SUBJECT'S LAST NAME: BELANGER

DATE/TIME OF TEST: 01/28/2017 16:05:19

Diagnostics Check		OK	15:59
Air Blank		0.000	15:59
Control Test		0.082	16:00
Air Blank		0.000	16:00
Subject Sample	# 1	0.000	16:01
Breath Volume		3.101	16:01
Air Blank		0.000	16:01
Air Blank		0.000	16:03
Subject Sample	# 2	0.000	16:03
Breath Volume		3.269	16:03
Air Blank		0.000	16:04
Control Test		0.080	16:04
Air Blank		0.000	16:05
Diagnostics Check		OK	16:05

Data Download Date/Time: 02/27/2017 13:36

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006168

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAY,ERICK,J
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: DAY,ERICK
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 01/20/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 18:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:43

SUBJECT'S LAST NAME: PARDO

DATE/TIME OF TEST: 01/29/2017 19:21:16

Diagnostics Check		OK	19:14
Air Blank		0.000	19:14
Control Test		0.082	19:15
Air Blank		0.000	19:15
Subject Sample	# 1	0.208	19:16
Breath Volume		2.308	19:16
Air Blank		0.000	19:17
Air Blank		0.000	19:18
Subject Sample	# 2	0.209	19:19
Breath Volume		2.226	19:19
Air Blank		0.000	19:20
Control Test		0.081	19:20
Air Blank		0.000	19:21
Diagnostics Check		OK	19:21

Data Download Date/Time: 02/27/2017 13:36

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006168

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANGERVIL,CLAUDY,P
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: ANGERVIL,CLAUDY
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 01/20/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:43

SUBJECT'S LAST NAME: CONLEY

DATE/TIME OF TEST: 01/30/2017 21:10:19

Diagnostics Check		OK	21:04
Air Blank		0.000	21:04
Control Test		0.082	21:04
Air Blank		0.000	21:05
Subject Sample	# 1	0.000	21:06
Breath Volume		1.921	21:06
Air Blank		0.000	21:06
Air Blank		0.000	21:08
Subject Sample	# 2	0.000	21:08
Breath Volume		2.082	21:08
Air Blank		0.000	21:09
Control Test		0.080	21:09
Air Blank		0.000	21:10
Diagnostics Check		OK	21:10

Data Download Date/Time: 02/27/2017 13:36

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006168

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELLAS,DILLON,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: BELLAS,DILLON
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 01/20/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:23

SUBJECT'S LAST NAME: DEWEY

DATE/TIME OF TEST: 01/31/2017 02:50:21

Diagnostics Check		OK	02:43
Air Blank		0.000	02:43
Control Test		0.082	02:44
Air Blank		0.000	02:44
Subject Sample	# 1	0.092	02:45
Breath Volume		1.753	02:45
Air Blank		0.000	02:46
Air Blank		0.000	02:47
Subject Sample	# 2	0.087	02:48
Breath Volume		1.554	02:48
Air Blank		0.000	02:49
Control Test		0.082	02:49
Air Blank		0.000	02:50
Diagnostics Check		OK	02:50

Data Download Date/Time: 02/27/2017 13:36

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006168

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 01/20/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 05:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:57

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 01/31/2017 06:25:53

Diagnostics Check		OK	06:19
Air Blank		0.000	06:20
Control Test		0.080	06:20
Air Blank		0.000	06:20
Subject Sample	# 1	0.200	06:21
Breath Volume		2.847	06:21
Air Blank		0.000	06:22
Air Blank		0.000	06:23
Subject Sample	# 2	0.203	06:24
Breath Volume		1.238	06:24
Air Blank		0.000	06:24
Control Test		0.079	06:25
Air Blank		0.000	06:25
Diagnostics Check		OK	06:25

Data Download Date/Time: 02/27/2017 13:36

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006168

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELLAS,DILLON,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: WISE,EUGENE
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 01/20/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 15:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:22

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 01/31/2017 17:54:58

Diagnostics Check		OK	17:48
Air Blank		0.000	17:48
Control Test		0.082	17:49
Air Blank		0.000	17:49
Subject Sample	# 1	0.105	17:50
Breath Volume		2.261	17:50
Air Blank		0.000	17:50
Air Blank		0.000	17:52
Subject Sample	# 2	0.103	17:53
Breath Volume		1.816	17:53
Air Blank		0.000	17:54
Control Test		0.081	17:54
Air Blank		0.000	17:54
Diagnostics Check		OK	17:54

Data Download Date/Time: 02/27/2017 13:36

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.