

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001275

INSTRUMENT REGISTERED TO: HIALEAH GARDENS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOROTIZA,JESUS,
BREATH TEST OPERATOR AGENCY: Hialeah Gardens PD
ARRESTING OFFICER: DORISMOND,WISNER
ARRESTING OFFICER AGENCY: Hialeah Gardens PD
LAST AGENCY INSPECTION: 02/16/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: DIAZ MORA

DATE/TIME OF TEST: 03/18/2017 01:41:58

Diagnostics Check	OK	01:39	
Air Blank	0.000	01:39	
Control Test	0.077	01:39	
Air Blank	0.000	01:40	
Subject Sample	# 1	01:40	Subject Test Refused
Breath Volume	0.000	01:40	
Air Blank	0.000	01:41	
Control Test	0.077	01:41	
Air Blank	0.000	01:41	
Diagnostics Check	OK	01:41	

Data Download Date/Time: 03/29/2017 13:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001275

INSTRUMENT REGISTERED TO: HIALEAH GARDENS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERNANDEZ,LISSETTE,C
BREATH TEST OPERATOR AGENCY: Medley PD
ARRESTING OFFICER: FERNANDEZ,LISSETTE
ARRESTING OFFICER AGENCY: Medley PD
LAST AGENCY INSPECTION: 02/16/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 05:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:58

SUBJECT'S LAST NAME: ORTIZ

DATE/TIME OF TEST: 03/23/2017 07:33:56

Diagnostics Check	OK	07:27
Air Blank	0.000	07:27
Control Test	0.079	07:28
Air Blank	0.000	07:28
Subject Sample # 1	0.176	07:29
Breath Volume	2.558	07:29
Air Blank	0.000	07:30
Air Blank	0.000	07:31
Subject Sample # 2	0.165	07:32
Breath Volume	3.484	07:32
Air Blank	0.000	07:33
Control Test	0.079	07:33
Air Blank	0.000	07:33
Diagnostics Check	OK	07:33

Data Download Date/Time: 03/29/2017 13:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KUNCAS,ANDREW,J
BREATH TEST OPERATOR AGENCY: Bal Harbour Vil PD
ARRESTING OFFICER: NAVAS,DAVID
ARRESTING OFFICER AGENCY: Bal Harbour Vil PD
LAST AGENCY INSPECTION: 03/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:07

SUBJECT'S LAST NAME: SOTO

DATE/TIME OF TEST: 03/18/2017 03:37:35

Diagnostics Check	OK	03:28
Air Blank	0.000	03:28
Control Test	0.080	03:29
Air Blank	0.000	03:29
Subject Sample # 1	0.127	03:31
Breath Volume	1.625	03:31
Air Blank	0.000	03:32
Air Blank	0.000	03:33
Subject Sample # 2	0.124	03:35
Breath Volume	1.183	03:35
Air Blank	0.000	03:36
Control Test	0.079	03:36
Air Blank	0.000	03:37
Diagnostics Check	OK	03:37

Data Download Date/Time: 04/10/2017 11:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 03/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:44

SUBJECT'S LAST NAME: BOSTICK

DATE/TIME OF TEST: 03/23/2017 00:23:52

Diagnostics Check	OK	00:08	
Air Blank	0.000	00:08	
Control Test	0.080	00:09	
Air Blank	0.000	00:09	
Subject Sample	# 1 0.146	00:11	
Breath Volume	1.585	00:11	
Air Blank	0.000	00:12	
Air Blank	0.000	00:13	
Subject Sample	# 2 0.125	00:17	Slope Not Level
Breath Volume	1.339	00:17	
Air Blank	0.000	00:17	
Air Blank	0.000	00:19	
Subject Sample	# 3	00:22	Subject Test Refused
Breath Volume	0.089	00:22	
Air Blank	0.000	00:22	
Control Test	0.080	00:23	
Air Blank	0.000	00:23	
Diagnostics Check	OK	00:23	

Data Download Date/Time: 04/10/2017 11:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JANVIER,FRITZ,G
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: JANVIER,FRITZ
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 03/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:49

SUBJECT'S LAST NAME: SEZER

DATE/TIME OF TEST: 03/29/2017 03:24:44

Diagnostics Check	OK	03:13	
Air Blank	0.000	03:13	
Control Test	0.080	03:14	
Air Blank	0.000	03:14	
Subject Sample	# 1 0.096	03:17	Volume Not Met
Breath Volume	0.101	03:17	
Air Blank	0.000	03:18	
Air Blank	0.000	03:20	
Subject Sample	# 2	03:23	Subject Test Refused
Breath Volume	0.648	03:23	
Air Blank	0.000	03:23	
Control Test	0.082	03:24	
Air Blank	0.000	03:24	
Diagnostics Check	OK	03:24	

Data Download Date/Time: 04/10/2017 11:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,STEVEN,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: ADAMS,STEVEN
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 03/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: MILANES-VALDES

DATE/TIME OF TEST: 03/11/2017 00:50:01

Diagnostics Check	OK	00:47	
Air Blank	0.000	00:47	
Control Test	0.079	00:47	
Air Blank	0.000	00:48	
Subject Sample	# 1	00:48	Subject Test Refused
Breath Volume	0.000	00:48	
Air Blank	0.000	00:49	
Control Test	0.079	00:49	
Air Blank	0.000	00:49	
Diagnostics Check	OK	00:50	

Data Download Date/Time: 04/10/2017 11:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,STEVEN,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: ADAMS,STEVEN
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 03/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: FAIRELL

DATE/TIME OF TEST: 03/11/2017 23:52:40

Diagnostics Check	OK	23:49	
Air Blank	0.000	23:50	
Control Test	0.079	23:50	
Air Blank	0.000	23:51	
Subject Sample	# 1	23:51	Subject Test Refused
Breath Volume	0.000	23:51	
Air Blank	0.000	23:51	
Control Test	0.079	23:52	
Air Blank	0.000	23:52	
Diagnostics Check	OK	23:52	

Data Download Date/Time: 04/10/2017 11:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,FRANK,E
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: GEBBIA,SALVATORE
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 03/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:51

SUBJECT'S LAST NAME: MYERS

DATE/TIME OF TEST: 03/17/2017 23:22:33

Diagnostics Check	OK	23:14
Air Blank	0.000	23:14
Control Test	0.079	23:15
Air Blank	0.000	23:15
Subject Sample # 1	0.241	23:16
Breath Volume	1.968	23:16
Air Blank	0.000	23:17
Air Blank	0.000	23:19
Subject Sample # 2	0.227	23:20
Breath Volume	2.121	23:20
Air Blank	0.000	23:21
Control Test	0.075	23:21
Air Blank	0.000	23:22
Diagnostics Check	OK	23:22

Data Download Date/Time: 04/10/2017 11:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY, SHANON, E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: EANES, J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:38

SUBJECT'S LAST NAME: FIELDS

DATE/TIME OF TEST: 03/01/2017 00:05:29

Diagnostics Check	OK	23:59
Air Blank	0.000	23:59
Control Test	0.080	23:59
Air Blank	0.000	00:00
Subject Sample # 1	0.237	00:01
Breath Volume	2.292	00:01
Air Blank	0.000	00:01
Air Blank	0.000	00:03
Subject Sample # 2	0.238	00:03
Breath Volume	1.449	00:03
Air Blank	0.000	00:04
Control Test	0.079	00:04
Air Blank	0.000	00:05
Diagnostics Check	OK	00:05

Data Download Date/Time: 03/15/2017 10:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THIGPEN,J,T
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MARRERO,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 15:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:41

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 03/01/2017 17:11:01

Diagnostics Check	OK	17:04
Air Blank	0.000	17:04
Control Test	0.080	17:05
Air Blank	0.000	17:05
Subject Sample # 1	0.151	17:06
Breath Volume	1.179	17:06
Air Blank	0.000	17:07
Air Blank	0.000	17:09
Subject Sample # 2	0.152	17:09
Breath Volume	1.414	17:09
Air Blank	0.000	17:10
Control Test	0.079	17:10
Air Blank	0.000	17:10
Diagnostics Check	OK	17:11

Data Download Date/Time: 03/15/2017 10:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THIGPEN,J,T
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BYRNE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 17:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:11

SUBJECT'S LAST NAME: PORTER

DATE/TIME OF TEST: 03/01/2017 18:40:03

Diagnostics Check	OK	18:33
Air Blank	0.000	18:34
Control Test	0.080	18:34
Air Blank	0.000	18:34
Subject Sample # 1	0.225	18:35
Breath Volume	3.425	18:35
Air Blank	0.000	18:36
Air Blank	0.000	18:37
Subject Sample # 2	0.209	18:38
Breath Volume	2.578	18:38
Air Blank	0.000	18:39
Control Test	0.079	18:39
Air Blank	0.000	18:39
Diagnostics Check	OK	18:40

Data Download Date/Time: 03/15/2017 10:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY, SHANON, E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FARLEY, J
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 02/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:07

SUBJECT'S LAST NAME: ISAACS

DATE/TIME OF TEST: 03/01/2017 23:34:02

Diagnostics Check	OK	23:27
Air Blank	0.000	23:28
Control Test	0.080	23:28
Air Blank	0.000	23:29
Subject Sample # 1	0.000	23:29
Breath Volume	4.429	23:29
Air Blank	0.000	23:30
Air Blank	0.000	23:32
Subject Sample # 2	0.000	23:32
Breath Volume	2.242	23:32
Air Blank	0.000	23:33
Control Test	0.080	23:33
Air Blank	0.000	23:33
Diagnostics Check	OK	23:34

Data Download Date/Time: 03/15/2017 10:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY, SHANON, E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: EANES, J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:04

SUBJECT'S LAST NAME: SEERAM

DATE/TIME OF TEST: 03/02/2017 02:28:00

Diagnostics Check	OK	02:25	
Air Blank	0.000	02:25	
Control Test	0.080	02:25	
Air Blank	0.000	02:26	
Subject Sample	# 1	02:26	Subject Test Refused
Breath Volume	0.000	02:26	
Air Blank	0.000	02:27	
Control Test	0.080	02:27	
Air Blank	0.000	02:27	
Diagnostics Check	OK	02:28	

Data Download Date/Time: 03/15/2017 10:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY, SHANON, E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BOWEN, J
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 02/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:16

SUBJECT'S LAST NAME: SILVA

DATE/TIME OF TEST: 03/02/2017 04:43:59

Diagnostics Check	OK	04:37
Air Blank	0.000	04:38
Control Test	0.080	04:38
Air Blank	0.000	04:38
Subject Sample # 1	0.000	04:39
Breath Volume	2.773	04:39
Air Blank	0.000	04:40
Air Blank	0.000	04:42
Subject Sample # 2	0.000	04:42
Breath Volume	2.261	04:42
Air Blank	0.000	04:43
Control Test	0.080	04:43
Air Blank	0.000	04:43
Diagnostics Check	OK	04:43

Data Download Date/Time: 03/15/2017 10:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY, SHANON, E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BAIS
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: GRASSMAN

DATE/TIME OF TEST: 03/02/2017 23:48:53

Diagnostics Check	OK	23:41
Air Blank	0.000	23:41
Control Test	0.081	23:42
Air Blank	0.000	23:42
Subject Sample # 1	0.125	23:43
Breath Volume	1.437	23:43
Air Blank	0.000	23:44
Air Blank	0.000	23:46
Subject Sample # 2	0.132	23:47
Breath Volume	1.144	23:47
Air Blank	0.000	23:47
Control Test	0.081	23:48
Air Blank	0.000	23:48
Diagnostics Check	OK	23:48

Data Download Date/Time: 03/15/2017 10:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY, SHANON, E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FARLEY, J
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 02/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: NUGENT

DATE/TIME OF TEST: 03/03/2017 02:11:42

Diagnostics Check	OK	02:05
Air Blank	0.000	02:06
Control Test	0.081	02:06
Air Blank	0.000	02:06
Subject Sample # 1	0.186	02:07
Breath Volume	2.421	02:07
Air Blank	0.000	02:07
Air Blank	0.000	02:09
Subject Sample # 2	0.179	02:10
Breath Volume	2.445	02:10
Air Blank	0.000	02:10
Control Test	0.081	02:11
Air Blank	0.000	02:11
Diagnostics Check	OK	02:11

Data Download Date/Time: 03/15/2017 10:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY, SHANON, E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LANE, C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:53

SUBJECT'S LAST NAME: WRIGHT

DATE/TIME OF TEST: 03/03/2017 04:19:18

Diagnostics Check	OK	04:13
Air Blank	0.000	04:13
Control Test	0.081	04:13
Air Blank	0.000	04:14
Subject Sample # 1	0.216	04:14
Breath Volume	2.390	04:14
Air Blank	0.000	04:15
Air Blank	0.000	04:17
Subject Sample # 2	0.208	04:17
Breath Volume	1.191	04:17
Air Blank	0.000	04:18
Control Test	0.080	04:18
Air Blank	0.000	04:19
Diagnostics Check	OK	04:19

Data Download Date/Time: 03/15/2017 10:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THIGPEN,J,T
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: GARCIA,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 17:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:35

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 03/05/2017 19:03:49

Diagnostics Check	OK	18:55
Air Blank	0.000	18:56
Control Test	0.081	18:56
Air Blank	0.000	18:57
Subject Sample # 1	0.205	18:58
Breath Volume	2.488	18:58
Air Blank	0.000	18:59
Air Blank	0.000	19:01
Subject Sample # 2	0.208	19:02
Breath Volume	2.605	19:02
Air Blank	0.000	19:02
Control Test	0.080	19:03
Air Blank	0.000	19:03
Diagnostics Check	OK	19:03

Data Download Date/Time: 03/15/2017 10:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THIGPEN,J,T
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: STANLEY,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:55

SUBJECT'S LAST NAME: MATEO

DATE/TIME OF TEST: 03/05/2017 21:27:22

Diagnostics Check	OK	21:19
Air Blank	0.000	21:20
Control Test	0.080	21:20
Air Blank	0.000	21:21
Subject Sample # 1	0.278	21:22
Breath Volume	3.316	21:22
Air Blank	0.000	21:23
Air Blank	0.000	21:25
Subject Sample # 2	0.267	21:25
Breath Volume	2.339	21:25
Air Blank	0.000	21:26
Control Test	0.079	21:26
Air Blank	0.000	21:27
Diagnostics Check	OK	21:27

Data Download Date/Time: 03/15/2017 10:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THIGPEN,J,T
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SAMUELS,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:44

SUBJECT'S LAST NAME: WASHINGTON

DATE/TIME OF TEST: 03/05/2017 22:12:43

Diagnostics Check	OK	22:06
Air Blank	0.000	22:06
Control Test	0.080	22:06
Air Blank	0.000	22:07
Subject Sample # 1	0.156	22:08
Breath Volume	2.929	22:08
Air Blank	0.000	22:08
Air Blank	0.000	22:10
Subject Sample # 2	0.162	22:11
Breath Volume	2.562	22:11
Air Blank	0.000	22:11
Control Test	0.080	22:12
Air Blank	0.000	22:12
Diagnostics Check	OK	22:12

Data Download Date/Time: 03/15/2017 10:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CAMACHO,K
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 02/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: LANIER

DATE/TIME OF TEST: 03/06/2017 03:29:03

Diagnostics Check	OK	03:26	
Air Blank	0.000	03:26	
Control Test	0.081	03:27	
Air Blank	0.000	03:27	
Subject Sample	# 1	03:27	Subject Test Refused
Breath Volume	0.000	03:27	
Air Blank	0.000	03:28	
Control Test	0.080	03:28	
Air Blank	0.000	03:28	
Diagnostics Check	OK	03:29	

Data Download Date/Time: 03/15/2017 10:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: THOMPSON,M
ARRESTING OFFICER AGENCY: Atlantic Beach PD
LAST AGENCY INSPECTION: 02/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:12

SUBJECT'S LAST NAME: NELSON

DATE/TIME OF TEST: 03/06/2017 04:38:51

Diagnostics Check	OK	04:32
Air Blank	0.000	04:32
Control Test	0.080	04:33
Air Blank	0.000	04:33
Subject Sample # 1	0.194	04:34
Breath Volume	1.621	04:34
Air Blank	0.000	04:34
Air Blank	0.000	04:36
Subject Sample # 2	0.206	04:37
Breath Volume	2.328	04:37
Air Blank	0.000	04:37
Control Test	0.080	04:38
Air Blank	0.000	04:38
Diagnostics Check	OK	04:38

Data Download Date/Time: 03/15/2017 10:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FARLEY,J
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 02/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:23

SUBJECT'S LAST NAME: SCOTT

DATE/TIME OF TEST: 03/07/2017 02:49:32

Diagnostics Check	OK	02:43
Air Blank	0.000	02:43
Control Test	0.080	02:43
Air Blank	0.000	02:44
Subject Sample # 1	0.078	02:45
Breath Volume	2.710	02:45
Air Blank	0.000	02:45
Air Blank	0.000	02:47
Subject Sample # 2	0.080	02:48
Breath Volume	2.179	02:48
Air Blank	0.000	02:48
Control Test	0.080	02:48
Air Blank	0.000	02:49
Diagnostics Check	OK	02:49

Data Download Date/Time: 03/15/2017 10:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PENNINGTON, EMILY, J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CAMACHO, K
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 02/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:32

SUBJECT'S LAST NAME: HUTTON

DATE/TIME OF TEST: 03/08/2017 02:58:59

Diagnostics Check	OK	02:52
Air Blank	0.000	02:53
Control Test	0.080	02:53
Air Blank	0.000	02:53
Subject Sample # 1	0.096	02:54
Breath Volume	2.425	02:54
Air Blank	0.000	02:54
Air Blank	0.000	02:56
Subject Sample # 2	0.100	02:57
Breath Volume	1.761	02:57
Air Blank	0.000	02:58
Control Test	0.080	02:58
Air Blank	0.000	02:58
Diagnostics Check	OK	02:58

Data Download Date/Time: 03/15/2017 10:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: GARCIA,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: ALSHMAOUN

DATE/TIME OF TEST: 03/09/2017 02:01:12

Diagnostics Check	OK	01:58	
Air Blank	0.000	01:58	
Control Test	0.080	01:59	
Air Blank	0.000	01:59	
Subject Sample	# 1	01:59	Subject Test Refused
Breath Volume	0.000	01:59	
Air Blank	0.000	02:00	
Control Test	0.080	02:00	
Air Blank	0.000	02:01	
Diagnostics Check	OK	02:01	

Data Download Date/Time: 03/15/2017 10:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY, SHANON, E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BAIS
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: BUDZASH

DATE/TIME OF TEST: 03/10/2017 02:48:57

Diagnostics Check	OK	02:45	
Air Blank	0.000	02:46	
Control Test	0.080	02:46	
Air Blank	0.000	02:46	
Subject Sample	# 1	02:47	Subject Test Refused
Breath Volume	0.000	02:47	
Air Blank	0.000	02:48	
Control Test	0.080	02:48	
Air Blank	0.000	02:48	
Diagnostics Check	OK	02:48	

Data Download Date/Time: 03/15/2017 10:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MURRIELL,E,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BRYANT,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: BLANTON

DATE/TIME OF TEST: 03/10/2017 21:21:38

Diagnostics Check	OK	21:13
Air Blank	0.000	21:14
Control Test	0.079	21:14
Air Blank	0.000	21:15
Subject Sample # 1	0.236	21:17
Breath Volume	1.914	21:17
Air Blank	0.000	21:17
Air Blank	0.000	21:19
Subject Sample # 2	0.250	21:20
Breath Volume	2.277	21:20
Air Blank	0.000	21:20
Control Test	0.079	21:21
Air Blank	0.000	21:21
Diagnostics Check	OK	21:21

Data Download Date/Time: 03/15/2017 10:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY, SHANON, E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FARLEY, J
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 02/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:37

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 03/12/2017 00:04:35

Diagnostics Check	OK	23:57
Air Blank	0.000	23:57
Control Test	0.080	23:57
Air Blank	0.000	23:58
Subject Sample # 2	0.017	23:59
Breath Volume	2.113	23:59
Air Blank	0.000	00:00
Air Blank	0.000	00:02
Subject Sample # 1	0.016	00:03
Breath Volume	2.453	00:03
Air Blank	0.000	00:03
Control Test	0.080	00:04
Air Blank	0.000	00:04
Diagnostics Check	OK	00:04

Data Download Date/Time: 03/15/2017 10:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY,SHANON,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BYRNE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:59

SUBJECT'S LAST NAME: DANIELS

DATE/TIME OF TEST: 03/12/2017 01:29:53

Diagnostics Check	OK	01:20
Air Blank	0.000	01:20
Control Test	0.080	01:21
Air Blank	0.000	01:21
Subject Sample # 1	0.197	01:22
Breath Volume	2.515	01:22
Air Blank	0.000	01:23
Air Blank	0.000	01:24
Subject Sample # 2	0.175	01:25
Breath Volume	1.246	01:25
Air Blank	0.000	01:26
Air Blank	0.000	01:27
Subject Sample # 3	0.200	01:28
Breath Volume	2.941	01:28
Air Blank	0.000	01:28
Control Test	0.079	01:29
Air Blank	0.000	01:29
Diagnostics Check	OK	01:29

Data Download Date/Time: 03/15/2017 10:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY, SHANON, E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FARLEY, J
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 02/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:31

SUBJECT'S LAST NAME: SHOOK

DATE/TIME OF TEST: 03/12/2017 04:01:54

Diagnostics Check	OK	03:52
Air Blank	0.000	03:52
Control Test	0.080	03:53
Air Blank	0.000	03:53
Subject Sample # 1	0.135	03:55
Breath Volume	1.312	03:55
Air Blank	0.000	03:56
Air Blank	0.000	03:58
Subject Sample # 2	0.130	04:00
Breath Volume	1.519	04:00
Air Blank	0.000	04:01
Control Test	0.078	04:01
Air Blank	0.000	04:01
Diagnostics Check	OK	04:01

Data Download Date/Time: 03/15/2017 10:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY,SHANON,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MILLER,C
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 02/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:36

SUBJECT'S LAST NAME: CHANCEY

DATE/TIME OF TEST: 03/12/2017 05:06:29

Diagnostics Check	OK	04:57	
Air Blank	0.000	04:57	
Control Test	0.080	04:58	
Air Blank	0.000	04:58	
Subject Sample	# 1	05:01	No Sample Provided
Breath Volume	0.000	05:01	
Air Blank	0.000	05:02	
Air Blank	0.000	05:04	
Subject Sample	# 2	05:05	Subject Test Refused
Breath Volume	0.000	05:05	
Air Blank	0.000	05:05	
Control Test	0.080	05:05	
Air Blank	0.000	05:06	
Diagnostics Check	OK	05:06	

Data Download Date/Time: 03/15/2017 10:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORTAGUS,B,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BROWN,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 16:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:22

SUBJECT'S LAST NAME: ELLIOTT

DATE/TIME OF TEST: 03/12/2017 19:48:13

Diagnostics Check	OK	19:43	
Air Blank	0.000	19:43	
Control Test	0.081	19:43	
Air Blank	0.000	19:44	
Subject Sample	# 1	19:46	Interferent Detect
Breath Volume	2.484	19:46	
Air Blank	0.000	19:47	
Control Test	0.080	19:47	
Air Blank	0.000	19:48	
Diagnostics Check	OK	19:48	

Data Download Date/Time: 03/15/2017 10:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY, SHANON, E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BYRNE, J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:18

SUBJECT'S LAST NAME: POLITOWSKI

DATE/TIME OF TEST: 03/13/2017 01:48:04

Diagnostics Check	OK	01:40
Air Blank	0.000	01:40
Control Test	0.080	01:41
Air Blank	0.000	01:41
Subject Sample # 1	0.285	01:42
Breath Volume	1.769	01:42
Air Blank	0.000	01:42
Air Blank	0.000	01:44
Subject Sample # 2	0.270	01:46
Breath Volume	2.027	01:46
Air Blank	0.000	01:47
Control Test	0.080	01:47
Air Blank	0.000	01:47
Diagnostics Check	OK	01:48

Data Download Date/Time: 03/15/2017 10:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY, SHANON, E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BYRNE, J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:44

SUBJECT'S LAST NAME: AWM

DATE/TIME OF TEST: 03/14/2017 01:11:29

Diagnostics Check	OK	01:04
Air Blank	0.000	01:04
Control Test	0.080	01:05
Air Blank	0.000	01:05
Subject Sample # 1	0.195	01:06
Breath Volume	1.257	01:06
Air Blank	0.000	01:06
Air Blank	0.000	01:08
Subject Sample # 2	0.197	01:09
Breath Volume	1.542	01:09
Air Blank	0.000	01:10
Control Test	0.080	01:10
Air Blank	0.000	01:11
Diagnostics Check	OK	01:11

Data Download Date/Time: 03/15/2017 10:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERY,SHANON,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BAIS
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:51

SUBJECT'S LAST NAME: SANDERS

DATE/TIME OF TEST: 03/14/2017 03:15:50

Diagnostics Check	OK	03:12	
Air Blank	0.000	03:13	
Control Test	0.081	03:13	
Air Blank	0.000	03:13	
Subject Sample	# 1	03:14	Subject Test Refused
Breath Volume	0.000	03:14	
Air Blank	0.000	03:14	
Control Test	0.081	03:15	
Air Blank	0.000	03:15	
Diagnostics Check	OK	03:15	

Data Download Date/Time: 03/15/2017 10:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY, SHANON, E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MILLER, C
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 02/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:08

SUBJECT'S LAST NAME: ELSASSER

DATE/TIME OF TEST: 03/14/2017 04:34:57

Diagnostics Check	OK	04:28
Air Blank	0.000	04:29
Control Test	0.080	04:29
Air Blank	0.000	04:29
Subject Sample # 1	0.214	04:30
Breath Volume	1.550	04:30
Air Blank	0.000	04:31
Air Blank	0.000	04:32
Subject Sample # 2	0.206	04:33
Breath Volume	1.796	04:33
Air Blank	0.000	04:34
Control Test	0.080	04:34
Air Blank	0.000	04:34
Diagnostics Check	OK	04:34

Data Download Date/Time: 03/15/2017 10:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY, SHANON, E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DZAMKO, J
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 02/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:06

SUBJECT'S LAST NAME: LANE

DATE/TIME OF TEST: 03/14/2017 05:29:54

Diagnostics Check	OK	05:27	
Air Blank	0.000	05:27	
Control Test	0.080	05:27	
Air Blank	0.000	05:28	
Subject Sample	# 1	05:28	Subject Test Refused
Breath Volume	0.000	05:28	
Air Blank	0.000	05:29	
Control Test	0.081	05:29	
Air Blank	0.000	05:29	
Diagnostics Check	OK	05:29	

Data Download Date/Time: 03/15/2017 10:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAUSEY,JOHN,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: STANLEY,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: YOUNT

DATE/TIME OF TEST: 03/14/2017 21:54:26

Diagnostics Check	OK	21:51	
Air Blank	0.000	21:52	
Control Test	0.081	21:52	
Air Blank	0.000	21:52	
Subject Sample	# 1	21:53	Subject Test Refused
Breath Volume	0.000	21:53	
Air Blank	0.000	21:53	
Control Test	0.080	21:53	
Air Blank	0.000	21:54	
Diagnostics Check	OK	21:54	

Data Download Date/Time: 03/15/2017 10:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAYLOR,TIMOTHY,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BROWN,T.A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:40

SUBJECT'S LAST NAME: RAINS

DATE/TIME OF TEST: 03/18/2017 21:04:55

Diagnostics Check	OK	21:01	
Air Blank	0.000	21:02	
Control Test	0.081	21:02	
Air Blank	0.000	21:03	
Subject Sample	# 1	21:03	Subject Test Refused
Breath Volume	0.000	21:03	
Air Blank	0.000	21:04	
Control Test	0.081	21:04	
Air Blank	0.000	21:04	
Diagnostics Check	OK	21:04	

Data Download Date/Time: 03/31/2017 11:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAYLOR,TIMOTHY,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: EANES,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: KLUZ

DATE/TIME OF TEST: 03/18/2017 22:07:50

Diagnostics Check	OK	22:01
Air Blank	0.000	22:01
Control Test	0.080	22:02
Air Blank	0.000	22:02
Subject Sample # 1	0.173	22:03
Breath Volume	3.531	22:03
Air Blank	0.000	22:03
Air Blank	0.000	22:05
Subject Sample # 2	0.163	22:06
Breath Volume	2.773	22:06
Air Blank	0.000	22:06
Control Test	0.080	22:07
Air Blank	0.000	22:07
Diagnostics Check	OK	22:07

Data Download Date/Time: 03/31/2017 11:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: EANES,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:03

SUBJECT'S LAST NAME: ROBLERO-GONZALEZ

DATE/TIME OF TEST: 03/19/2017 00:30:34

Diagnostics Check	OK	00:24
Air Blank	0.000	00:24
Control Test	0.080	00:24
Air Blank	0.000	00:25
Subject Sample # 1	0.085	00:26
Breath Volume	2.828	00:26
Air Blank	0.000	00:26
Air Blank	0.000	00:28
Subject Sample # 2	0.084	00:29
Breath Volume	2.300	00:29
Air Blank	0.000	00:29
Control Test	0.080	00:29
Air Blank	0.000	00:30
Diagnostics Check	OK	00:30

Data Download Date/Time: 03/31/2017 11:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: KERSHAW,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: NIEDZWIECKI

DATE/TIME OF TEST: 03/19/2017 02:17:24

Diagnostics Check	OK	02:11
Air Blank	0.000	02:11
Control Test	0.080	02:12
Air Blank	0.000	02:12
Subject Sample # 1	0.130	02:13
Breath Volume	1.960	02:13
Air Blank	0.000	02:13
Air Blank	0.000	02:15
Subject Sample # 2	0.132	02:15
Breath Volume	1.468	02:15
Air Blank	0.000	02:16
Control Test	0.080	02:16
Air Blank	0.000	02:17
Diagnostics Check	OK	02:17

Data Download Date/Time: 03/31/2017 11:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BYRNE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: SHERMAN

DATE/TIME OF TEST: 03/19/2017 03:12:13

Diagnostics Check	OK	03:06
Air Blank	0.000	03:06
Control Test	0.080	03:06
Air Blank	0.000	03:07
Subject Sample # 1	0.180	03:07
Breath Volume	2.015	03:07
Air Blank	0.000	03:08
Air Blank	0.000	03:10
Subject Sample # 2	0.164	03:10
Breath Volume	1.699	03:10
Air Blank	0.000	03:11
Control Test	0.079	03:11
Air Blank	0.000	03:12
Diagnostics Check	OK	03:12

Data Download Date/Time: 03/31/2017 11:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: KERSHAW,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:56

SUBJECT'S LAST NAME: HENDRIX

DATE/TIME OF TEST: 03/19/2017 05:22:26

Diagnostics Check	OK	05:16
Air Blank	0.000	05:16
Control Test	0.080	05:17
Air Blank	0.000	05:17
Subject Sample # 1	0.000	05:18
Breath Volume	1.265	05:18
Air Blank	0.000	05:18
Air Blank	0.000	05:20
Subject Sample # 2	0.000	05:21
Breath Volume	2.910	05:21
Air Blank	0.000	05:21
Control Test	0.080	05:21
Air Blank	0.000	05:22
Diagnostics Check	OK	05:22

Data Download Date/Time: 03/31/2017 11:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAYLOR,TIMOTHY,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MEYERS,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 16:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:40

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 03/19/2017 17:08:49

Diagnostics Check	OK	17:02
Air Blank	0.000	17:02
Control Test	0.081	17:03
Air Blank	0.000	17:03
Subject Sample # 1	0.174	17:04
Breath Volume	1.160	17:04
Air Blank	0.000	17:04
Air Blank	0.000	17:06
Subject Sample # 2	0.178	17:07
Breath Volume	2.585	17:07
Air Blank	0.000	17:07
Control Test	0.081	17:08
Air Blank	0.000	17:08
Diagnostics Check	OK	17:08

Data Download Date/Time: 03/31/2017 11:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAYLOR,TIMOTHY,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: EANES,J,L
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 18:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:01

SUBJECT'S LAST NAME: HERRERA

DATE/TIME OF TEST: 03/19/2017 20:29:05

Diagnostics Check	OK	20:21
Air Blank	0.000	20:22
Control Test	0.081	20:22
Air Blank	0.000	20:22
Subject Sample # 1	0.142	20:24
Breath Volume	3.027	20:24
Air Blank	0.000	20:25
Air Blank	0.000	20:26
Subject Sample # 2	0.142	20:27
Breath Volume	1.296	20:27
Air Blank	0.000	20:28
Control Test	0.079	20:28
Air Blank	0.000	20:29
Diagnostics Check	OK	20:29

Data Download Date/Time: 03/31/2017 11:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRUETT,RICHARD,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FISSEL,M
ARRESTING OFFICER AGENCY: Atlantic Beach PD
LAST AGENCY INSPECTION: 03/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: YARBROUGH

DATE/TIME OF TEST: 03/20/2017 03:36:47

Diagnostics Check	OK	03:29
Air Blank	0.000	03:30
Control Test	0.081	03:30
Air Blank	0.000	03:31
Subject Sample # 1	0.164	03:32
Breath Volume	1.535	03:32
Air Blank	0.000	03:33
Air Blank	0.000	03:34
Subject Sample # 2	0.163	03:35
Breath Volume	1.988	03:35
Air Blank	0.000	03:35
Control Test	0.080	03:36
Air Blank	0.000	03:36
Diagnostics Check	OK	03:36

Data Download Date/Time: 03/31/2017 11:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRUETT,RICHARD,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BYRNE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: VANBRUNT

DATE/TIME OF TEST: 03/20/2017 04:30:47

Diagnostics Check	OK	04:20	
Air Blank	0.000	04:21	
Control Test	0.081	04:21	
Air Blank	0.000	04:22	
Subject Sample	# 1 0.232	04:24	
Breath Volume	2.164	04:24	
Air Blank	0.000	04:25	
Air Blank	0.000	04:27	
Subject Sample	# 2	04:29	Subject Test Refused
Breath Volume	0.000	04:29	
Air Blank	0.000	04:29	
Control Test	0.079	04:30	
Air Blank	0.000	04:30	
Diagnostics Check	OK	04:30	

Data Download Date/Time: 03/31/2017 11:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRUETT,RICHARD,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: VEREEN,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 05:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:53

SUBJECT'S LAST NAME: SWEENEY

DATE/TIME OF TEST: 03/20/2017 06:20:29

Diagnostics Check	OK	06:14
Air Blank	0.000	06:14
Control Test	0.081	06:14
Air Blank	0.000	06:15
Subject Sample # 1	0.179	06:15
Breath Volume	1.582	06:15
Air Blank	0.000	06:16
Air Blank	0.000	06:18
Subject Sample # 2	0.194	06:18
Breath Volume	2.699	06:18
Air Blank	0.000	06:19
Control Test	0.080	06:19
Air Blank	0.000	06:20
Diagnostics Check	OK	06:20

Data Download Date/Time: 03/31/2017 11:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MURRIELL,E,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WILLIAMS,D
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 03/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:36

SUBJECT'S LAST NAME: VOLACHACK

DATE/TIME OF TEST: 03/22/2017 01:05:20

Diagnostics Check	OK	00:57
Air Blank	0.000	00:57
Control Test	0.080	00:58
Air Blank	0.000	00:58
Subject Sample # 1	0.074	01:00
Breath Volume	1.320	01:00
Air Blank	0.000	01:00
Air Blank	0.000	01:02
Subject Sample # 2	0.071	01:03
Breath Volume	1.265	01:03
Air Blank	0.000	01:04
Control Test	0.080	01:04
Air Blank	0.000	01:05
Diagnostics Check	OK	01:05

Data Download Date/Time: 03/31/2017 11:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MURRIELL,E,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HOOVER,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:28

SUBJECT'S LAST NAME: SALTMARSH

DATE/TIME OF TEST: 03/23/2017 22:51:45

Diagnostics Check	OK	22:48	
Air Blank	0.000	22:49	
Control Test	0.079	22:49	
Air Blank	0.000	22:50	
Subject Sample	# 1	22:50	Subject Test Refused
Breath Volume	0.000	22:50	
Air Blank	0.000	22:50	
Control Test	0.080	22:51	
Air Blank	0.000	22:51	
Diagnostics Check	OK	22:51	

Data Download Date/Time: 03/31/2017 11:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCOTT,J,A
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: HEALEY,G
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: WARD

DATE/TIME OF TEST: 03/24/2017 03:53:36

Diagnostics Check	OK	03:45
Air Blank	0.000	03:46
Control Test	0.080	03:46
Air Blank	0.000	03:47
Subject Sample # 1	0.095	03:49
Breath Volume	1.648	03:49
Air Blank	0.000	03:49
Air Blank	0.000	03:51
Subject Sample # 2	0.100	03:52
Breath Volume	1.714	03:52
Air Blank	0.000	03:52
Control Test	0.080	03:53
Air Blank	0.000	03:53
Diagnostics Check	OK	03:53

Data Download Date/Time: 03/31/2017 11:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRUETT,RICHARD,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FARLEY,J
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 03/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:49

SUBJECT'S LAST NAME: STRONG

DATE/TIME OF TEST: 03/25/2017 04:16:04

Diagnostics Check	OK	04:09
Air Blank	0.000	04:10
Control Test	0.080	04:10
Air Blank	0.000	04:10
Subject Sample # 1	0.219	04:11
Breath Volume	1.378	04:11
Air Blank	0.000	04:12
Air Blank	0.000	04:13
Subject Sample # 2	0.221	04:14
Breath Volume	2.855	04:14
Air Blank	0.000	04:15
Control Test	0.080	04:15
Air Blank	0.000	04:15
Diagnostics Check	OK	04:16

Data Download Date/Time: 03/31/2017 11:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAYLOR,TIMOTHY,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ONORATO,B
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 03/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 15:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:46

SUBJECT'S LAST NAME: ASHMAN

DATE/TIME OF TEST: 03/25/2017 20:14:06

Diagnostics Check	OK	20:07
Air Blank	0.000	20:07
Control Test	0.080	20:08
Air Blank	0.000	20:08
Subject Sample # 1	0.191	20:09
Breath Volume	2.539	20:09
Air Blank	0.000	20:10
Air Blank	0.000	20:11
Subject Sample # 2	0.189	20:12
Breath Volume	1.988	20:12
Air Blank	0.000	20:13
Control Test	0.079	20:13
Air Blank	0.000	20:14
Diagnostics Check	OK	20:14

Data Download Date/Time: 03/31/2017 11:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAYLOR,TIMOTHY,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: GARCIA,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 03/25/2017 21:53:58

Diagnostics Check	OK	21:51	
Air Blank	0.000	21:51	
Control Test	0.080	21:51	
Air Blank	0.000	21:52	
Subject Sample	# 1	21:52	Subject Test Refused
Breath Volume	0.000	21:52	
Air Blank	0.000	21:53	
Control Test	0.080	21:53	
Air Blank	0.000	21:53	
Diagnostics Check	OK	21:53	

Data Download Date/Time: 03/31/2017 11:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MURRIELL,EUGENE,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BROWN,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/15/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:42

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 03/30/2017 00:05:28

Diagnostics Check	OK	00:02	
Air Blank	0.000	00:03	
Control Test	0.080	00:03	
Air Blank	0.000	00:03	
Subject Sample	# 1	00:04	Subject Test Refused
Breath Volume	0.000	00:04	
Air Blank	0.000	00:04	
Control Test	0.080	00:04	
Air Blank	0.000	00:05	
Diagnostics Check	OK	00:05	

Data Download Date/Time: 03/31/2017 11:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MURRIELL,E,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: EANES,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: LEAKE

DATE/TIME OF TEST: 03/30/2017 02:42:33

Diagnostics Check	OK	02:35
Air Blank	0.000	02:36
Control Test	0.080	02:36
Air Blank	0.000	02:36
Subject Sample # 1	0.144	02:37
Breath Volume	1.417	02:37
Air Blank	0.000	02:38
Air Blank	0.000	02:40
Subject Sample # 2	0.138	02:41
Breath Volume	1.496	02:41
Air Blank	0.000	02:41
Control Test	0.080	02:41
Air Blank	0.000	02:42
Diagnostics Check	OK	02:42

Data Download Date/Time: 03/31/2017 11:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MURRIELL,E,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MARRERO,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:12

SUBJECT'S LAST NAME: BRANTLEY

DATE/TIME OF TEST: 03/30/2017 04:38:50

Diagnostics Check	OK	04:32
Air Blank	0.000	04:32
Control Test	0.080	04:33
Air Blank	0.000	04:33
Subject Sample # 1	0.155	04:34
Breath Volume	2.046	04:34
Air Blank	0.000	04:34
Air Blank	0.000	04:36
Subject Sample # 2	0.157	04:37
Breath Volume	1.964	04:37
Air Blank	0.000	04:37
Control Test	0.079	04:38
Air Blank	0.000	04:38
Diagnostics Check	OK	04:38

Data Download Date/Time: 03/31/2017 11:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTER,JENIFER,R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MARRERO,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: CARTER

DATE/TIME OF TEST: 03/30/2017 23:36:57

Diagnostics Check	OK	23:30
Air Blank	0.000	23:31
Control Test	0.080	23:31
Air Blank	0.000	23:32
Subject Sample # 1	0.194	23:32
Breath Volume	1.949	23:32
Air Blank	0.000	23:33
Air Blank	0.000	23:35
Subject Sample # 2	0.182	23:35
Breath Volume	2.015	23:35
Air Blank	0.000	23:36
Control Test	0.080	23:36
Air Blank	0.000	23:36
Diagnostics Check	OK	23:36

Data Download Date/Time: 03/31/2017 11:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCOTT,J,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BACOLOR,A
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 03/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: DWYER

DATE/TIME OF TEST: 03/31/2017 03:46:35

Diagnostics Check	OK	03:40
Air Blank	0.000	03:41
Control Test	0.080	03:41
Air Blank	0.000	03:41
Subject Sample # 1	0.087	03:42
Breath Volume	3.406	03:42
Air Blank	0.000	03:42
Air Blank	0.000	03:44
Subject Sample # 2	0.091	03:45
Breath Volume	2.765	03:45
Air Blank	0.000	03:45
Control Test	0.081	03:45
Air Blank	0.000	03:46
Diagnostics Check	OK	03:46

Data Download Date/Time: 03/31/2017 11:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THIGPEN,J,T
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FILBERT,E
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 03/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: ENCINO-HUERTA

DATE/TIME OF TEST: 03/31/2017 22:37:44

Diagnostics Check	OK	22:31
Air Blank	0.000	22:31
Control Test	0.080	22:32
Air Blank	0.000	22:32
Subject Sample # 1	0.217	22:33
Breath Volume	2.167	22:33
Air Blank	0.000	22:33
Air Blank	0.000	22:35
Subject Sample # 2	0.213	22:36
Breath Volume	2.179	22:36
Air Blank	0.000	22:36
Control Test	0.080	22:37
Air Blank	0.000	22:37
Diagnostics Check	OK	22:37

Data Download Date/Time: 04/13/2017 13:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOSTON,DONALD,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FUTCH,G
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:59

SUBJECT'S LAST NAME: MARTORANO

DATE/TIME OF TEST: 03/10/2017 23:26:27

Diagnostics Check	OK	23:20
Air Blank	0.000	23:20
Control Test	0.080	23:20
Air Blank	0.000	23:21
Subject Sample # 1	0.172	23:21
Breath Volume	3.945	23:21
Air Blank	0.000	23:22
Air Blank	0.000	23:24
Subject Sample # 2	0.156	23:24
Breath Volume	1.703	23:24
Air Blank	0.000	23:25
Control Test	0.079	23:25
Air Blank	0.000	23:26
Diagnostics Check	OK	23:26

Data Download Date/Time: 03/15/2017 10:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOSTON,DONALD,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FARELY,J
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 02/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:49

SUBJECT'S LAST NAME: FRAZIER

DATE/TIME OF TEST: 03/11/2017 01:16:35

Diagnostics Check	OK	01:09
Air Blank	0.000	01:10
Control Test	0.079	01:10
Air Blank	0.000	01:11
Subject Sample # 1	0.175	01:11
Breath Volume	3.789	01:11
Air Blank	0.000	01:12
Air Blank	0.000	01:14
Subject Sample # 2	0.178	01:15
Breath Volume	3.644	01:15
Air Blank	0.000	01:15
Control Test	0.080	01:15
Air Blank	0.000	01:16
Diagnostics Check	OK	01:16

Data Download Date/Time: 03/15/2017 10:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOSTON,DONALD,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WILLIAMS,D
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 02/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: SAAVEDRA GARCIA

DATE/TIME OF TEST: 03/13/2017 01:47:26

Diagnostics Check	OK	01:40
Air Blank	0.000	01:41
Control Test	0.080	01:41
Air Blank	0.000	01:42
Subject Sample # 1	0.203	01:42
Breath Volume	1.929	01:42
Air Blank	0.000	01:43
Air Blank	0.000	01:45
Subject Sample # 2	0.211	01:45
Breath Volume	1.136	01:45
Air Blank	0.000	01:46
Control Test	0.080	01:46
Air Blank	0.000	01:47
Diagnostics Check	OK	01:47

Data Download Date/Time: 03/15/2017 10:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JAMES, ALFONSO,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: COFFEY, E
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 07:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:16

SUBJECT'S LAST NAME: VASQUEZ RODRIGUEZ

DATE/TIME OF TEST: 03/20/2017 09:44:23

Diagnostics Check	OK	09:37
Air Blank	0.000	09:37
Control Test	0.081	09:37
Air Blank	0.000	09:38
Subject Sample # 1	0.200	09:39
Breath Volume	1.316	09:39
Air Blank	0.000	09:40
Air Blank	0.000	09:42
Subject Sample # 2	0.195	09:42
Breath Volume	1.589	09:42
Air Blank	0.000	09:43
Control Test	0.080	09:43
Air Blank	0.000	09:44
Diagnostics Check	OK	09:44

Data Download Date/Time: 03/31/2017 11:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOSTON,DONALD,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: VANHOF,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:19

SUBJECT'S LAST NAME: STELMA

DATE/TIME OF TEST: 03/23/2017 04:46:37

Diagnostics Check	OK	04:39
Air Blank	0.000	04:40
Control Test	0.080	04:40
Air Blank	0.000	04:41
Subject Sample # 1	0.105	04:41
Breath Volume	1.214	04:41
Air Blank	0.000	04:42
Air Blank	0.000	04:44
Subject Sample # 2	0.117	04:45
Breath Volume	3.917	04:45
Air Blank	0.000	04:45
Control Test	0.081	04:46
Air Blank	0.000	04:46
Diagnostics Check	OK	04:46

Data Download Date/Time: 03/31/2017 11:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001282

INSTRUMENT REGISTERED TO: UNIVERSITY OF FL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAIN II,KENNETH,A
BREATH TEST OPERATOR AGENCY: U of FI PD
ARRESTING OFFICER: BAIN II,KENNETH
ARRESTING OFFICER AGENCY: U of FI PD
LAST AGENCY INSPECTION: 02/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A2
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:58

SUBJECT'S LAST NAME: REAGLE

DATE/TIME OF TEST: 03/03/2017 03:35:32

Diagnostics Check	OK	03:23	
Air Blank	0.000	03:23	
Control Test	0.079	03:24	
Air Blank	0.000	03:24	
Subject Sample	# 1 0.188	03:27	Slope Not Level
Breath Volume	1.214	03:27	
Air Blank	0.000	03:28	
Air Blank	0.000	03:30	
Subject Sample	# 2 0.185	03:31	
Breath Volume	1.191	03:31	
Air Blank	0.000	03:31	
Air Blank	0.000	03:33	
Subject Sample	# 3 0.200	03:34	
Breath Volume	1.183	03:34	
Air Blank	0.000	03:34	
Control Test	0.079	03:34	
Air Blank	0.000	03:35	
Diagnostics Check	OK	03:35	

Data Download Date/Time: 03/31/2017 17:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001282

INSTRUMENT REGISTERED TO: UNIVERSITY OF FL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BECKHAM,MARCUS,W
BREATH TEST OPERATOR AGENCY: U of FL PD
ARRESTING OFFICER: BECKHAM,MARCUS
ARRESTING OFFICER AGENCY: U of FL PD
LAST AGENCY INSPECTION: 02/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A2
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:18

SUBJECT'S LAST NAME: SOUTHLAND

DATE/TIME OF TEST: 03/03/2017 04:03:29

Diagnostics Check	OK	03:57
Air Blank	0.000	03:57
Control Test	0.077	03:58
Air Blank	0.000	03:58
Subject Sample # 1	0.000	03:59
Breath Volume	1.285	03:59
Air Blank	0.000	03:59
Air Blank	0.000	04:01
Subject Sample # 2	0.000	04:02
Breath Volume	1.242	04:02
Air Blank	0.000	04:02
Control Test	0.078	04:02
Air Blank	0.000	04:03
Diagnostics Check	OK	04:03

Data Download Date/Time: 03/31/2017 17:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001282

INSTRUMENT REGISTERED TO: UNIVERSITY OF FL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEWART,DILLON,J
BREATH TEST OPERATOR AGENCY: U of FL PD
ARRESTING OFFICER: GORDON,COREY
ARRESTING OFFICER AGENCY: U of FL PD
LAST AGENCY INSPECTION: 02/27/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 24515080A2
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:52

SUBJECT'S LAST NAME: FIGUEROA

DATE/TIME OF TEST: 03/05/2017 02:35:35

Diagnostics Check	OK	02:21	
Air Blank	0.000	02:21	
Control Test	0.077	02:21	
Air Blank	0.000	02:22	
Subject Sample	# 1	02:25	No Sample Provided
Breath Volume	0.000	02:25	
Air Blank	0.000	02:25	
Air Blank	0.000	02:27	
Subject Sample	# 2	0.218	02:28
Breath Volume	1.878	02:28	
Air Blank	0.000	02:29	
Air Blank	0.000	02:30	
Subject Sample	# 3	0.210	02:33
Breath Volume	1.316	02:33	
Air Blank	0.000	02:34	
Control Test	0.079	02:34	
Air Blank	0.000	02:35	
Diagnostics Check	OK	02:35	

Data Download Date/Time: 03/31/2017 17:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001282

INSTRUMENT REGISTERED TO: UNIVERSITY OF FL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLWAY, CHAD, S
BREATH TEST OPERATOR AGENCY: U of FL PD
ARRESTING OFFICER: HOLWAY, CHAD
ARRESTING OFFICER AGENCY: U of FL PD
LAST AGENCY INSPECTION: 02/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A2
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:02

SUBJECT'S LAST NAME: GRGURIC

DATE/TIME OF TEST: 03/10/2017 01:46:43

Diagnostics Check	OK	01:35	
Air Blank	0.000	01:35	
Control Test	0.077	01:36	
Air Blank	0.000	01:36	
Subject Sample	# 1	01:39	No Sample Provided
Breath Volume	0.000	01:39	
Air Blank	0.000	01:40	
Air Blank	0.000	01:42	
Subject Sample	# 2	01:45	No Sample Provided
Breath Volume	0.000	01:45	
Air Blank	0.000	01:45	
Control Test	0.079	01:46	
Air Blank	0.000	01:46	
Diagnostics Check	OK	01:46	

Data Download Date/Time: 03/31/2017 17:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001282

INSTRUMENT REGISTERED TO: UNIVERSITY OF FL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BECKHAM,MARCUS,W
BREATH TEST OPERATOR AGENCY: U of FL PD
ARRESTING OFFICER: BAIN,KENNETH
ARRESTING OFFICER AGENCY: U of FL PD
LAST AGENCY INSPECTION: 02/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A2
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:16

SUBJECT'S LAST NAME: VENUTO

DATE/TIME OF TEST: 03/11/2017 03:47:15

Diagnostics Check	OK	03:40
Air Blank	0.000	03:40
Control Test	0.079	03:40
Air Blank	0.000	03:41
Subject Sample # 1	0.093	03:42
Breath Volume	1.792	03:42
Air Blank	0.000	03:42
Air Blank	0.000	03:44
Subject Sample # 2	0.087	03:45
Breath Volume	1.933	03:45
Air Blank	0.000	03:46
Control Test	0.079	03:46
Air Blank	0.000	03:47
Diagnostics Check	OK	03:47

Data Download Date/Time: 03/31/2017 17:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001282

INSTRUMENT REGISTERED TO: UNIVERSITY OF FL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLWAY, CHAD, S
BREATH TEST OPERATOR AGENCY: U of FL PD
ARRESTING OFFICER: BEGUIRISTAIN, VICTOR
ARRESTING OFFICER AGENCY: U of FL PD
LAST AGENCY INSPECTION: 02/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A2
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:01

SUBJECT'S LAST NAME: STOCKMAN

DATE/TIME OF TEST: 03/15/2017 02:28:14

Diagnostics Check	OK	02:21
Air Blank	0.000	02:21
Control Test	0.079	02:22
Air Blank	0.000	02:22
Subject Sample # 1	0.159	02:23
Breath Volume	2.195	02:23
Air Blank	0.000	02:24
Air Blank	0.000	02:26
Subject Sample # 2	0.150	02:26
Breath Volume	1.937	02:26
Air Blank	0.000	02:27
Control Test	0.079	02:27
Air Blank	0.000	02:28
Diagnostics Check	OK	02:28

Data Download Date/Time: 03/31/2017 17:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001282

INSTRUMENT REGISTERED TO: UNIVERSITY OF FL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLWAY, CHAD, S
BREATH TEST OPERATOR AGENCY: U of FL PD
ARRESTING OFFICER: ZARATE, JESSICA
ARRESTING OFFICER AGENCY: U of FL PD
LAST AGENCY INSPECTION: 02/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A2
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:21

SUBJECT'S LAST NAME: MANDANI

DATE/TIME OF TEST: 03/18/2017 01:57:47

Diagnostics Check	OK	01:48
Air Blank	0.000	01:49
Control Test	0.079	01:49
Air Blank	0.000	01:50
Subject Sample # 1	0.175	01:50
Breath Volume	1.199	01:50
Air Blank	0.000	01:51
Air Blank	0.000	01:53
Subject Sample # 2	0.176	01:56
Breath Volume	1.316	01:56
Air Blank	0.000	01:56
Control Test	0.080	01:57
Air Blank	0.000	01:57
Diagnostics Check	OK	01:57

Data Download Date/Time: 03/31/2017 17:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001282

INSTRUMENT REGISTERED TO: UNIVERSITY OF FL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAIN II,KENNETH,A
BREATH TEST OPERATOR AGENCY: U of FI PD
ARRESTING OFFICER: BECKHAM,MARCUS
ARRESTING OFFICER AGENCY: U of FI PD
LAST AGENCY INSPECTION: 02/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A2
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:06

SUBJECT'S LAST NAME: CASARIEGO

DATE/TIME OF TEST: 03/22/2017 02:35:53

Diagnostics Check	OK	02:27
Air Blank	0.000	02:27
Control Test	0.079	02:28
Air Blank	0.000	02:28
Subject Sample # 1	0.098	02:30
Breath Volume	1.195	02:30
Air Blank	0.000	02:31
Air Blank	0.000	02:32
Subject Sample # 2	0.096	02:34
Breath Volume	1.519	02:34
Air Blank	0.000	02:34
Control Test	0.079	02:35
Air Blank	0.000	02:35
Diagnostics Check	OK	02:35

Data Download Date/Time: 03/31/2017 17:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001282

INSTRUMENT REGISTERED TO: UNIVERSITY OF FL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BECKHAM,MARCUS,W
BREATH TEST OPERATOR AGENCY: U of FI PD
ARRESTING OFFICER: SASSER,WILLIAM
ARRESTING OFFICER AGENCY: U of FI PD
LAST AGENCY INSPECTION: 02/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A2
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: COLLINS

DATE/TIME OF TEST: 03/25/2017 03:00:31

Diagnostics Check	OK	02:51
Air Blank	0.000	02:52
Control Test	0.079	02:52
Air Blank	0.000	02:53
Subject Sample # 1	0.169	02:55
Breath Volume	1.585	02:55
Air Blank	0.000	02:56
Air Blank	0.000	02:58
Subject Sample # 2	0.175	02:58
Breath Volume	1.882	02:58
Air Blank	0.000	02:59
Control Test	0.077	02:59
Air Blank	0.000	03:00
Diagnostics Check	OK	03:00

Data Download Date/Time: 03/31/2017 17:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001282

INSTRUMENT REGISTERED TO: UNIVERSITY OF FL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BECKHAM,MARCUS,W
BREATH TEST OPERATOR AGENCY: U of FL PD
ARRESTING OFFICER: BAIN,KENNETH
ARRESTING OFFICER AGENCY: U of FL PD
LAST AGENCY INSPECTION: 02/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A2
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: ALQARNI

DATE/TIME OF TEST: 03/25/2017 03:14:09

Diagnostics Check	OK	03:06
Air Blank	0.000	03:06
Control Test	0.077	03:07
Air Blank	0.000	03:07
Subject Sample # 1	0.088	03:09
Breath Volume	1.390	03:09
Air Blank	0.000	03:10
Air Blank	0.000	03:12
Subject Sample # 2	0.090	03:12
Breath Volume	2.257	03:12
Air Blank	0.000	03:13
Control Test	0.078	03:13
Air Blank	0.000	03:14
Diagnostics Check	OK	03:14

Data Download Date/Time: 03/31/2017 17:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001285

INSTRUMENT REGISTERED TO: GILCHRIST COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOOVER,JESSE,W
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: HOOVER,JESSE
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 02/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 259585
DRY GAS EXPIRY DATE: 01/12/2019
VIOLATION CODE: DUI
ARREST TIME: 03:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: GORDON

DATE/TIME OF TEST: 03/12/2017 04:19:20

Diagnostics Check	OK	04:07	
Air Blank	0.000	04:07	
Control Test	0.077	04:08	
Air Blank	0.000	04:08	
Subject Sample	# 1 0.242	04:11	Volume Not Met
Breath Volume	1.035	04:11	
Air Blank	0.000	04:12	
Air Blank	0.000	04:14	
Subject Sample	# 2 0.235	04:17	Volume Not Met
Breath Volume	0.898	04:17	
Air Blank	0.000	04:18	
Control Test	0.073	04:18	Control Outside Tolerance
Air Blank	0.000	04:19	

Data Download Date/Time: 03/16/2017 01:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001285

INSTRUMENT REGISTERED TO: GILCHRIST COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOOVER,JESSE,W
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: HOOVER,JESSE
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 02/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 259585
DRY GAS EXPIRY DATE: 01/12/2019
VIOLATION CODE: DUI
ARREST TIME: 03:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: GORDON

DATE/TIME OF TEST: 03/12/2017 04:29:03

Diagnostics Check	OK	04:21	
Air Blank	0.000	04:21	
Control Test	0.076	04:22	
Air Blank	0.000	04:22	
Subject Sample	# 1 0.231	04:24	
Breath Volume	1.160	04:24	
Air Blank	0.000	04:25	
Air Blank	0.000	04:27	
Subject Sample	# 2 0.247	04:27	
Breath Volume	1.308	04:27	
Air Blank	0.000	04:28	
Control Test	0.074	04:28	Control Outside Tolerance
Air Blank	0.000	04:29	

Data Download Date/Time: 03/16/2017 01:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001285

INSTRUMENT REGISTERED TO: GILCHRIST COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THWEATT,CAMERON,J
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: THWEATT,CAMERON
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 03/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 259585
DRY GAS EXPIRY DATE: 01/12/2019
VIOLATION CODE: DUI
ARREST TIME: 23:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:17

SUBJECT'S LAST NAME: HERNANDEZ-DIAZ

DATE/TIME OF TEST: 03/31/2017 23:50:57

Diagnostics Check	OK	23:43
Air Blank	0.000	23:44
Control Test	0.077	23:44
Air Blank	0.000	23:44
Subject Sample # 1	0.204	23:46
Breath Volume	3.011	23:46
Air Blank	0.000	23:47
Air Blank	0.000	23:48
Subject Sample # 2	0.199	23:49
Breath Volume	2.984	23:49
Air Blank	0.000	23:50
Control Test	0.077	23:50
Air Blank	0.000	23:50
Diagnostics Check	OK	23:50

Data Download Date/Time: 04/28/2017 00:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001287

INSTRUMENT REGISTERED TO: BAKER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FROST,KEVIN,W
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: FROST,KEVIN
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 02/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:23

SUBJECT'S LAST NAME: REDMOND

DATE/TIME OF TEST: 03/03/2017 21:46:33

Diagnostics Check	OK	21:43	
Air Blank	0.000	21:44	
Control Test	0.083	21:44	
Air Blank	0.000	21:45	
Subject Sample	# 1	21:45	Subject Test Refused
Breath Volume	0.000	21:45	
Air Blank	0.000	21:45	
Control Test	0.083	21:45	
Air Blank	0.000	21:46	
Diagnostics Check	OK	21:46	

Data Download Date/Time: 03/31/2017 20:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001287

INSTRUMENT REGISTERED TO: BAKER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HILLIARD,KARI,A
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: BURNS,JUSTIN
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 02/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 18:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:00

SUBJECT'S LAST NAME: AVELLA

DATE/TIME OF TEST: 03/10/2017 18:29:09

Diagnostics Check	OK	18:22
Air Blank	0.000	18:23
Control Test	0.082	18:23
Air Blank	0.000	18:24
Subject Sample # 1	0.174	18:24
Breath Volume	1.640	18:24
Air Blank	0.000	18:25
Air Blank	0.000	18:27
Subject Sample # 2	0.169	18:27
Breath Volume	1.714	18:27
Air Blank	0.000	18:28
Control Test	0.080	18:28
Air Blank	0.000	18:29
Diagnostics Check	OK	18:29

Data Download Date/Time: 03/31/2017 20:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001289

INSTRUMENT REGISTERED TO: UNION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HILLIARD,KARI,A
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: BOATRIGHT,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 02/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1824329
DRY GAS EXPIRY DATE: 03/09/2018
VIOLATION CODE: DUI
ARREST TIME: 21:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:37

SUBJECT'S LAST NAME: KING

DATE/TIME OF TEST: 03/12/2017 23:03:53

Diagnostics Check	OK	22:57
Air Blank	0.000	22:58
Control Test	0.081	22:58
Air Blank	0.000	22:59
Subject Sample # 1	0.000	22:59
Breath Volume	1.660	22:59
Air Blank	0.000	23:00
Air Blank	0.000	23:02
Subject Sample # 2	0.000	23:02
Breath Volume	1.890	23:02
Air Blank	0.000	23:02
Control Test	0.077	23:03
Air Blank	0.000	23:03
Diagnostics Check	OK	23:03

Data Download Date/Time: 03/31/2017 15:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001289

INSTRUMENT REGISTERED TO: UNION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JANKOWSKI,JAMES,F
BREATH TEST OPERATOR AGENCY: Union County SO
ARRESTING OFFICER: JANKOWSKI,JAMES
ARRESTING OFFICER AGENCY: Union County SO
LAST AGENCY INSPECTION: 02/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1824329
DRY GAS EXPIRY DATE: 03/09/2018
VIOLATION CODE: DUI
ARREST TIME: 01:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:52

SUBJECT'S LAST NAME: HICKS

DATE/TIME OF TEST: 03/19/2017 02:19:36

Diagnostics Check	OK	02:13
Air Blank	0.000	02:13
Control Test	0.080	02:14
Air Blank	0.000	02:14
Subject Sample # 1	0.136	02:15
Breath Volume	4.175	02:15
Air Blank	0.000	02:15
Air Blank	0.000	02:17
Subject Sample # 2	0.131	02:18
Breath Volume	2.890	02:18
Air Blank	0.000	02:18
Control Test	0.081	02:19
Air Blank	0.000	02:19
Diagnostics Check	OK	02:19

Data Download Date/Time: 03/31/2017 15:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001289

INSTRUMENT REGISTERED TO: UNION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,CHARLES,C
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: ANDREWS,SCOTT
ARRESTING OFFICER AGENCY: Union County SO
LAST AGENCY INSPECTION: 02/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1824329
DRY GAS EXPIRY DATE: 03/09/2018
VIOLATION CODE: DUI
ARREST TIME: 20:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:06

SUBJECT'S LAST NAME: CURTIS

DATE/TIME OF TEST: 03/25/2017 21:32:47

Diagnostics Check	OK	21:26
Air Blank	0.000	21:26
Control Test	0.079	21:27
Air Blank	0.000	21:27
Subject Sample # 1	0.371	21:28
Breath Volume	2.179	21:28
Air Blank	0.000	21:28
Air Blank	0.000	21:30
Subject Sample # 2	0.369	21:31
Breath Volume	2.031	21:31
Air Blank	0.000	21:31
Control Test	0.079	21:32
Air Blank	0.000	21:32
Diagnostics Check	OK	21:32

Data Download Date/Time: 03/31/2017 15:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001291

INSTRUMENT REGISTERED TO: BRADFORD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HULLENDER,ROBERT,S
BREATH TEST OPERATOR AGENCY: Bradford County SO
ARRESTING OFFICER: MCREE,GARETT
ARRESTING OFFICER AGENCY: Bradford County SO
LAST AGENCY INSPECTION: 02/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG619605
DRY GAS EXPIRY DATE: 07/14/2018
VIOLATION CODE: DUI
ARREST TIME: 15:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:20

SUBJECT'S LAST NAME: MOBLEY

DATE/TIME OF TEST: 03/09/2017 15:51:28

Diagnostics Check	OK	15:43	
Air Blank	0.000	15:43	
Control Test	0.080	15:43	
Air Blank	0.000	15:44	
Subject Sample	# 1	15:47	No Sample Provided
Breath Volume	0.000	15:47	
Air Blank	0.000	15:47	
Air Blank	0.000	15:49	
Subject Sample	# 2	15:50	Subject Test Refused
Breath Volume	0.000	15:50	
Air Blank	0.000	15:50	
Control Test	0.079	15:50	
Air Blank	0.000	15:51	
Diagnostics Check	OK	15:51	

Data Download Date/Time: 03/13/2017 17:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001291

INSTRUMENT REGISTERED TO: BRADFORD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHILLER,JOSEPH,C
BREATH TEST OPERATOR AGENCY: Bradford County SO
ARRESTING OFFICER: SHILLER,JOSEPH
ARRESTING OFFICER AGENCY: Bradford County SO
LAST AGENCY INSPECTION: 02/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG619605
DRY GAS EXPIRY DATE: 07/14/2018
VIOLATION CODE: DUI
ARREST TIME: 19:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:20

SUBJECT'S LAST NAME: BANKS

DATE/TIME OF TEST: 03/10/2017 20:46:26

Diagnostics Check	OK	20:40
Air Blank	0.000	20:40
Control Test	0.079	20:40
Air Blank	0.000	20:41
Subject Sample # 1	0.035	20:41
Breath Volume	2.503	20:41
Air Blank	0.000	20:42
Air Blank	0.000	20:44
Subject Sample # 2	0.033	20:44
Breath Volume	1.742	20:44
Air Blank	0.000	20:45
Control Test	0.078	20:45
Air Blank	0.000	20:46
Diagnostics Check	OK	20:46

Data Download Date/Time: 03/13/2017 17:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001291

INSTRUMENT REGISTERED TO: BRADFORD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,TERRELL,G
BREATH TEST OPERATOR AGENCY: Bradford County SO
ARRESTING OFFICER: WILLIAMS,TERRELL
ARRESTING OFFICER AGENCY: Bradford County SO
LAST AGENCY INSPECTION: 03/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG619605
DRY GAS EXPIRY DATE: 07/14/2018
VIOLATION CODE: DUI
ARREST TIME: 19:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:10

SUBJECT'S LAST NAME: MCINTOSH

DATE/TIME OF TEST: 03/15/2017 19:52:35

Diagnostics Check	OK	19:46
Air Blank	0.000	19:46
Control Test	0.081	19:47
Air Blank	0.000	19:47
Subject Sample # 1	0.000	19:48
Breath Volume	1.496	19:48
Air Blank	0.000	19:48
Air Blank	0.000	19:50
Subject Sample # 2	0.000	19:51
Breath Volume	1.515	19:51
Air Blank	0.000	19:51
Control Test	0.079	19:52
Air Blank	0.000	19:52
Diagnostics Check	OK	19:52

Data Download Date/Time: 04/28/2017 20:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001291

INSTRUMENT REGISTERED TO: BRADFORD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GORDON,RUSSELL,B
BREATH TEST OPERATOR AGENCY: Bradford County SO
ARRESTING OFFICER: GORDON,RUSSELL
ARRESTING OFFICER AGENCY: Bradford County SO
LAST AGENCY INSPECTION: 03/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG619605
DRY GAS EXPIRY DATE: 07/14/2018
VIOLATION CODE: DUI
ARREST TIME: 22:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:08

SUBJECT'S LAST NAME: REPLOGLE

DATE/TIME OF TEST: 03/26/2017 22:51:46

Diagnostics Check	OK	22:48	
Air Blank	0.000	22:49	
Control Test	0.078	22:49	
Air Blank	0.000	22:50	
Subject Sample	# 1	22:50	Subject Test Refused
Breath Volume	0.000	22:50	
Air Blank	0.000	22:50	
Control Test	0.078	22:51	
Air Blank	0.000	22:51	
Diagnostics Check	OK	22:51	

Data Download Date/Time: 04/28/2017 20:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001291

INSTRUMENT REGISTERED TO: BRADFORD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HULLENDER,ROBERT,S
BREATH TEST OPERATOR AGENCY: Bradford County SO
ARRESTING OFFICER: HULLENDER,ROBERT
ARRESTING OFFICER AGENCY: Bradford County SO
LAST AGENCY INSPECTION: 03/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG619605
DRY GAS EXPIRY DATE: 07/14/2018
VIOLATION CODE: DUI
ARREST TIME: 22:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:57

SUBJECT'S LAST NAME: WHEELER

DATE/TIME OF TEST: 03/31/2017 23:36:17

Diagnostics Check	OK	23:29
Air Blank	0.000	23:30
Control Test	0.078	23:30
Air Blank	0.000	23:30
Subject Sample # 1	0.113	23:31
Breath Volume	3.035	23:31
Air Blank	0.000	23:32
Air Blank	0.000	23:33
Subject Sample # 2	0.114	23:34
Breath Volume	2.593	23:34
Air Blank	0.000	23:35
Control Test	0.076	23:35
Air Blank	0.000	23:36
Diagnostics Check	OK	23:36

Data Download Date/Time: 04/28/2017 20:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001293

INSTRUMENT REGISTERED TO: WILLISTON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANDIS,BRYAN,S
BREATH TEST OPERATOR AGENCY: Williston PD
ARRESTING OFFICER: LONG,KATHLEEN
ARRESTING OFFICER AGENCY: Williston PD
LAST AGENCY INSPECTION: 02/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 00:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: CRIBBS

DATE/TIME OF TEST: 03/06/2017 01:36:59

Diagnostics Check	OK	01:30
Air Blank	0.000	01:31
Control Test	0.080	01:31
Air Blank	0.000	01:31
Subject Sample # 1	0.073	01:32
Breath Volume	1.507	01:32
Air Blank	0.000	01:33
Air Blank	0.000	01:35
Subject Sample # 2	0.071	01:35
Breath Volume	1.550	01:35
Air Blank	0.000	01:36
Control Test	0.080	01:36
Air Blank	0.000	01:36
Diagnostics Check	OK	01:36

Data Download Date/Time: 03/25/2017 23:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001293

INSTRUMENT REGISTERED TO: WILLISTON PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: WRIGHT,MARQUIS,A

BREATH TEST OPERATOR AGENCY: Williston PD

ARRESTING OFFICER: WRIGHT,MARQUIS

ARRESTING OFFICER AGENCY: Williston PD

LAST AGENCY INSPECTION: 02/27/2017

Date or Time Changed: No

DRY GAS STANDARD LOT #: 203498

DRY GAS EXPIRY DATE: 12/10/2018

VIOLATION CODE: DUI

ARREST TIME: 12:33

GENDER: Female

OBSERVATION PERIOD BEGAN: 13:17

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 03/09/2017 13:51:02

Diagnostics Check	OK	13:44
Air Blank	0.000	13:45
Control Test	0.080	13:45
Air Blank	0.000	13:46
Subject Sample # 1	0.000	13:46
Breath Volume	1.437	13:46
Air Blank	0.000	13:47
Air Blank	0.000	13:49
Subject Sample # 2	0.000	13:49
Breath Volume	2.304	13:49
Air Blank	0.000	13:50
Control Test	0.079	13:50
Air Blank	0.000	13:50
Diagnostics Check	OK	13:51

Data Download Date/Time: 03/25/2017 23:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001293

INSTRUMENT REGISTERED TO: WILLISTON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANDIS,BRYAN,S
BREATH TEST OPERATOR AGENCY: Williston PD
ARRESTING OFFICER: LANDIS,BRYAN
ARRESTING OFFICER AGENCY: Williston PD
LAST AGENCY INSPECTION: 02/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 01:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: DAVIES

DATE/TIME OF TEST: 03/10/2017 02:06:41

Diagnostics Check	OK	02:03	
Air Blank	0.000	02:03	
Control Test	0.079	02:04	
Air Blank	0.000	02:04	
Subject Sample	# 1	02:05	Subject Test Refused
Breath Volume	0.000	02:05	
Air Blank	0.000	02:05	
Control Test	0.080	02:06	
Air Blank	0.000	02:06	
Diagnostics Check	OK	02:06	

Data Download Date/Time: 03/25/2017 23:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001293

INSTRUMENT REGISTERED TO: WILLISTON PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MORRIS,FRED,A

BREATH TEST OPERATOR AGENCY: Williston PD

ARRESTING OFFICER: LONG,KATHLEEN

ARRESTING OFFICER AGENCY: Williston PD

LAST AGENCY INSPECTION: 02/27/2017

Date or Time Changed: No

DRY GAS STANDARD LOT #: 203498

DRY GAS EXPIRY DATE: 12/10/2018

VIOLATION CODE: DUI

ARREST TIME: 03:27

GENDER: Female

OBSERVATION PERIOD BEGAN: 04:40

SUBJECT'S LAST NAME: POTTER

DATE/TIME OF TEST: 03/18/2017 05:18:16

Diagnostics Check	OK	05:08
Air Blank	0.000	05:08
Control Test	0.079	05:08
Air Blank	0.000	05:09
Subject Sample # 1	0.173	05:11
Breath Volume	1.496	05:11
Air Blank	0.000	05:12
Air Blank	0.000	05:13
Subject Sample # 2	0.167	05:16
Breath Volume	2.042	05:16
Air Blank	0.000	05:17
Control Test	0.078	05:17
Air Blank	0.000	05:18
Diagnostics Check	OK	05:18

Data Download Date/Time: 03/25/2017 23:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VITALE,MICHAEL,L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: VITALE,MICHAEL
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 02/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG606103
DRY GAS EXPIRY DATE: 03/01/2018
VIOLATION CODE: DUI
ARREST TIME: 23:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:19

SUBJECT'S LAST NAME: KELLEY

DATE/TIME OF TEST: 03/01/2017 00:42:55

Diagnostics Check	OK	00:40	
Air Blank	0.000	00:40	
Control Test	0.079	00:40	
Air Blank	0.000	00:41	
Subject Sample	# 1	00:41	Subject Test Refused
Breath Volume	0.000	00:41	
Air Blank	0.000	00:42	
Control Test	0.080	00:42	
Air Blank	0.000	00:42	
Diagnostics Check	OK	00:42	

Data Download Date/Time: 03/30/2017 20:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VITALE,MICHAEL,L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: YOUNGHANSE,JOHN
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 02/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG606103
DRY GAS EXPIRY DATE: 03/01/2018
VIOLATION CODE: DUI
ARREST TIME: 23:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:28

SUBJECT'S LAST NAME: FORRISTER

DATE/TIME OF TEST: 03/01/2017 00:59:26

Diagnostics Check	OK	00:52
Air Blank	0.000	00:52
Control Test	0.079	00:52
Air Blank	0.000	00:53
Subject Sample # 1	0.213	00:55
Breath Volume	3.882	00:55
Air Blank	0.000	00:55
Air Blank	0.000	00:57
Subject Sample # 2	0.198	00:57
Breath Volume	3.375	00:57
Air Blank	0.000	00:58
Control Test	0.078	00:58
Air Blank	0.000	00:59
Diagnostics Check	OK	00:59

Data Download Date/Time: 03/30/2017 20:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VITALE,MICHAEL,L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: VITALE,MICHAEL
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 02/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629801
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 00:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:46

SUBJECT'S LAST NAME: KELLY

DATE/TIME OF TEST: 03/03/2017 02:14:26

Diagnostics Check	OK	02:06
Air Blank	0.000	02:07
Control Test	0.080	02:07
Air Blank	0.000	02:08
Subject Sample # 1	0.137	02:09
Breath Volume	1.445	02:09
Air Blank	0.000	02:10
Air Blank	0.000	02:11
Subject Sample # 2	0.143	02:12
Breath Volume	2.277	02:12
Air Blank	0.000	02:13
Control Test	0.080	02:13
Air Blank	0.000	02:14
Diagnostics Check	OK	02:14

Data Download Date/Time: 03/30/2017 20:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VITALE,MICHAEL,L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: VITALE,MICHAEL
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 02/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629801
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 23:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: CRUTHIRDS

DATE/TIME OF TEST: 03/04/2017 00:16:44

Diagnostics Check	OK	00:10
Air Blank	0.000	00:10
Control Test	0.080	00:11
Air Blank	0.000	00:11
Subject Sample # 1	0.173	00:12
Breath Volume	2.660	00:12
Air Blank	0.000	00:12
Air Blank	0.000	00:14
Subject Sample # 2	0.174	00:15
Breath Volume	3.226	00:15
Air Blank	0.000	00:15
Control Test	0.080	00:16
Air Blank	0.000	00:16
Diagnostics Check	OK	00:16

Data Download Date/Time: 03/30/2017 20:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TERHUNE,JEREMY,M
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: TERHUNE,JEREMY
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 02/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629801
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 03:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: EDWARDS

DATE/TIME OF TEST: 03/04/2017 04:13:14

Diagnostics Check	OK	04:05
Air Blank	0.000	04:05
Control Test	0.080	04:06
Air Blank	0.000	04:06
Subject Sample # 1	0.118	04:08
Breath Volume	1.218	04:08
Air Blank	0.000	04:08
Air Blank	0.000	04:10
Subject Sample # 2	0.119	04:11
Breath Volume	2.496	04:11
Air Blank	0.000	04:12
Control Test	0.080	04:12
Air Blank	0.000	04:13
Diagnostics Check	OK	04:13

Data Download Date/Time: 03/30/2017 20:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TUSLER,DANNY,K
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: BODNOVITS,JASON
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 02/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629801
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 04:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:50

SUBJECT'S LAST NAME: SALAS

DATE/TIME OF TEST: 03/04/2017 05:18:08

Diagnostics Check	OK	05:10
Air Blank	0.000	05:11
Control Test	0.080	05:11
Air Blank	0.000	05:12
Subject Sample # 1	0.161	05:13
Breath Volume	2.316	05:13
Air Blank	0.000	05:13
Air Blank	0.000	05:15
Subject Sample # 2	0.160	05:16
Breath Volume	2.277	05:16
Air Blank	0.000	05:17
Control Test	0.078	05:17
Air Blank	0.000	05:18
Diagnostics Check	OK	05:18

Data Download Date/Time: 03/30/2017 20:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VITALE,MICHAEL,L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: VITALE,MICHAEL
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 02/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629801
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 02:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: LASSITTER

DATE/TIME OF TEST: 03/05/2017 03:57:53

Diagnostics Check	OK	03:51
Air Blank	0.000	03:52
Control Test	0.078	03:52
Air Blank	0.000	03:53
Subject Sample # 1	0.152	03:53
Breath Volume	1.421	03:53
Air Blank	0.000	03:54
Air Blank	0.000	03:55
Subject Sample # 2	0.162	03:56
Breath Volume	1.394	03:56
Air Blank	0.000	03:56
Control Test	0.079	03:57
Air Blank	0.000	03:57
Diagnostics Check	OK	03:57

Data Download Date/Time: 03/30/2017 20:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHOTSCH,RYAN,A
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: SCHOTSCH,RYAN
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 02/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629801
DRY GAS EXPIRY DATE: 08/24/2018
VIOLATION CODE: DUI
ARREST TIME: 03:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 03/05/2017 04:39:06

Diagnostics Check	OK	04:28	
Air Blank	0.000	04:29	
Control Test	0.080	04:29	
Air Blank	0.000	04:29	
Subject Sample	# 1	04:33	No Sample Provided
Breath Volume	0.000	04:33	
Air Blank	0.000	04:33	
Air Blank	0.000	04:35	
Subject Sample	# 2	04:37	Subject Test Refused
Breath Volume	0.000	04:37	
Air Blank	0.000	04:38	
Control Test	0.079	04:38	
Air Blank	0.000	04:38	
Diagnostics Check	OK	04:39	

Data Download Date/Time: 03/30/2017 20:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VITALE,MICHAEL,L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: VITALE,MICHAEL
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 02/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629801
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 22:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: REEB

DATE/TIME OF TEST: 03/06/2017 00:17:19

Diagnostics Check	OK	00:11
Air Blank	0.000	00:11
Control Test	0.080	00:11
Air Blank	0.000	00:12
Subject Sample # 1	0.185	00:12
Breath Volume	4.445	00:12
Air Blank	0.000	00:13
Air Blank	0.000	00:15
Subject Sample # 2	0.170	00:15
Breath Volume	2.570	00:15
Air Blank	0.000	00:16
Control Test	0.079	00:16
Air Blank	0.000	00:17
Diagnostics Check	OK	00:17

Data Download Date/Time: 03/30/2017 20:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRIAN,ALTON,D
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: BERRIAN,ALTON
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 02/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629801
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 00:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: DENNISON

DATE/TIME OF TEST: 03/08/2017 01:42:47

Diagnostics Check	OK	01:41	
Air Blank	0.000	01:41	
Control Test	0.079	01:41	
Air Blank		01:42	RFI Detect
Air Blank	0.000	01:42	

Data Download Date/Time: 03/30/2017 20:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRIAN,ALTON,D
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: BERRIAN,ALTON
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 02/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629801
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 00:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: DENNISON

DATE/TIME OF TEST: 03/08/2017 01:52:51

Diagnostics Check	OK	01:46
Air Blank	0.000	01:46
Control Test	0.080	01:47
Air Blank	0.000	01:47
Subject Sample # 1	0.372	01:48
Breath Volume	1.691	01:48
Air Blank	0.000	01:48
Air Blank	0.000	01:50
Subject Sample # 2	0.378	01:51
Breath Volume	1.820	01:51
Air Blank	0.000	01:51
Control Test	0.078	01:52
Air Blank	0.000	01:52
Diagnostics Check	OK	01:52

Data Download Date/Time: 03/30/2017 20:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEINER,ZACHARY,A
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: STEINER,ZACHARY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 02/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629801
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 08:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:40

SUBJECT'S LAST NAME: HUNT

DATE/TIME OF TEST: 03/08/2017 10:04:36

Diagnostics Check	OK	10:01	
Air Blank	0.000	10:01	
Control Test	0.079	10:02	
Air Blank	0.000	10:02	
Subject Sample # 1	0.000	10:03	
Breath Volume	2.824	10:03	
Air Blank		10:03	Improper Sample
Air Blank	0.000	10:03	
Control Test	0.079	10:04	
Air Blank	0.000	10:04	
Diagnostics Check	OK	10:04	

Data Download Date/Time: 03/30/2017 20:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEINER,ZACHARY,A
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: STEINER,ZACHARY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 02/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629801
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 08:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:40

SUBJECT'S LAST NAME: HUNT

DATE/TIME OF TEST: 03/08/2017 10:11:58

Diagnostics Check	OK	10:06
Air Blank	0.000	10:06
Control Test	0.079	10:06
Air Blank	0.000	10:07
Subject Sample # 1	0.000	10:07
Breath Volume	3.269	10:07
Air Blank	0.000	10:08
Air Blank	0.000	10:10
Subject Sample # 2	0.000	10:10
Breath Volume	3.281	10:10
Air Blank	0.000	10:11
Control Test	0.079	10:11
Air Blank	0.000	10:11
Diagnostics Check	OK	10:11

Data Download Date/Time: 03/30/2017 20:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOHEN,DANIEL,A
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: KOHEN,DANIEL
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 02/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629801
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 22:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: ODILLON

DATE/TIME OF TEST: 03/09/2017 00:16:52

Diagnostics Check	OK	00:08
Air Blank	0.000	00:09
Control Test	0.078	00:09
Air Blank	0.000	00:09
Subject Sample # 1	0.109	00:12
Breath Volume	1.582	00:12
Air Blank	0.000	00:13
Air Blank	0.000	00:14
Subject Sample # 2	0.106	00:15
Breath Volume	1.875	00:15
Air Blank	0.000	00:15
Control Test	0.077	00:16
Air Blank	0.000	00:16
Diagnostics Check	OK	00:16

Data Download Date/Time: 03/30/2017 20:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TOLLIVER,JORDON,P
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: TOLLIVER,JORDON
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 02/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629801
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 20:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: BLANTON

DATE/TIME OF TEST: 03/10/2017 21:43:35

Diagnostics Check	OK	21:39	
Air Blank	0.000	21:40	
Control Test	0.080	21:40	
Air Blank	0.000	21:41	
Subject Sample	# 1	21:42	Subject Test Refused
Breath Volume	0.000	21:42	
Air Blank	0.000	21:42	
Control Test	0.079	21:43	
Air Blank	0.000	21:43	
Diagnostics Check	OK	21:43	

Data Download Date/Time: 03/30/2017 20:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLOYD,CHAD,M
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: FLOYD,CHAD
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 02/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629801
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: PROB/PAROL
ARREST TIME: 22:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 03/10/2017 22:53:43

Diagnostics Check	OK	22:47
Air Blank	0.000	22:48
Control Test	0.079	22:48
Air Blank	0.000	22:48
Subject Sample # 1	0.000	22:49
Breath Volume	3.042	22:49
Air Blank	0.000	22:49
Air Blank	0.000	22:51
Subject Sample # 2	0.000	22:52
Breath Volume	3.179	22:52
Air Blank	0.000	22:52
Control Test	0.079	22:53
Air Blank	0.000	22:53
Diagnostics Check	OK	22:53

Data Download Date/Time: 03/30/2017 20:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VITALE,MICHAEL,L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: VITALE,MICHAEL
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 02/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629801
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 03:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:08

SUBJECT'S LAST NAME: LEONARD

DATE/TIME OF TEST: 03/11/2017 04:35:46

Diagnostics Check	OK	04:28
Air Blank	0.000	04:29
Control Test	0.080	04:29
Air Blank	0.000	04:29
Subject Sample # 1	0.153	04:31
Breath Volume	1.472	04:31
Air Blank	0.000	04:32
Air Blank	0.000	04:33
Subject Sample # 2	0.158	04:34
Breath Volume	1.195	04:34
Air Blank	0.000	04:34
Control Test	0.079	04:35
Air Blank	0.000	04:35
Diagnostics Check	OK	04:35

Data Download Date/Time: 03/30/2017 20:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LADUE,ASHLEY,E
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: LADUE,ASHLEY
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 02/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629801
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: PROB/PAROL
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:00

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 03/11/2017 11:42:49

Diagnostics Check	OK	11:36
Air Blank	0.000	11:37
Control Test	0.080	11:37
Air Blank	0.000	11:38
Subject Sample # 1	0.000	11:38
Breath Volume	3.796	11:38
Air Blank	0.000	11:39
Air Blank	0.000	11:41
Subject Sample # 2	0.000	11:41
Breath Volume	3.941	11:41
Air Blank	0.000	11:41
Control Test	0.080	11:42
Air Blank	0.000	11:42
Diagnostics Check	OK	11:42

Data Download Date/Time: 03/30/2017 20:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALLACE,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: WALLACE,CHRISTOPHER
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 02/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629801
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 18:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:08

SUBJECT'S LAST NAME: KREVATAS

DATE/TIME OF TEST: 03/11/2017 19:56:31

Diagnostics Check	OK	19:50
Air Blank	0.000	19:50
Control Test	0.080	19:50
Air Blank	0.000	19:51
Subject Sample # 1	0.000	19:51
Breath Volume	1.707	19:51
Air Blank	0.000	19:52
Air Blank	0.000	19:54
Subject Sample # 2	0.000	19:55
Breath Volume	2.375	19:55
Air Blank	0.000	19:55
Control Test	0.079	19:55
Air Blank	0.000	19:56
Diagnostics Check	OK	19:56

Data Download Date/Time: 03/30/2017 20:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUICK,STEPHEN,W
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: QUICK,STEPHEN
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 02/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629801
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: OTHER
ARREST TIME: 23:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: HINES

DATE/TIME OF TEST: 03/15/2017 00:51:08

Diagnostics Check	OK	00:46	
Air Blank	0.000	00:46	
Control Test	0.079	00:46	
Air Blank	0.000	00:47	
Subject Sample	# 1	00:50	Sequence Aborted
Breath Volume	0.140	00:50	
Air Blank	0.000	00:51	

Data Download Date/Time: 03/30/2017 20:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WELFORD,DERIC,M
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: POND,CURTISS
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 02/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629801
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 17:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:20

SUBJECT'S LAST NAME: COLEMAN

DATE/TIME OF TEST: 03/16/2017 18:50:41

Diagnostics Check	OK	18:40	
Air Blank	0.000	18:41	
Control Test	0.080	18:41	
Air Blank	0.000	18:42	
Subject Sample	# 1 0.035	18:45	Volume Not Met
Breath Volume	0.597	18:45	
Air Blank	0.000	18:45	
Air Blank	0.000	18:47	
Subject Sample	# 2	18:49	Subject Test Refused
Breath Volume	0.527	18:49	
Air Blank	0.000	18:49	
Control Test	0.080	18:50	
Air Blank	0.000	18:50	
Diagnostics Check	OK	18:50	

Data Download Date/Time: 03/30/2017 20:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,JOHN,F
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: JOHNSON,JORDAN
ARRESTING OFFICER AGENCY: Milton PD
LAST AGENCY INSPECTION: 02/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629801
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:54

SUBJECT'S LAST NAME: BUTLER

DATE/TIME OF TEST: 03/17/2017 02:20:51

Diagnostics Check	OK	02:14
Air Blank	0.000	02:14
Control Test	0.080	02:15
Air Blank	0.000	02:15
Subject Sample # 1	0.158	02:16
Breath Volume	1.945	02:16
Air Blank	0.000	02:16
Air Blank	0.000	02:18
Subject Sample # 2	0.163	02:19
Breath Volume	2.531	02:19
Air Blank	0.000	02:19
Control Test	0.080	02:20
Air Blank	0.000	02:20
Diagnostics Check	OK	02:20

Data Download Date/Time: 03/30/2017 20:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPENCER, RAYMOND, L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: SPENCER, RAYMOND
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 02/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629801
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 17:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:05

SUBJECT'S LAST NAME: BOSWELL

DATE/TIME OF TEST: 03/17/2017 18:37:39

Diagnostics Check	OK	18:25	
Air Blank	0.000	18:26	
Control Test	0.079	18:26	
Air Blank	0.000	18:27	
Subject Sample	# 1	18:30	No Sample Provided
Breath Volume	0.000	18:30	
Air Blank	0.000	18:30	
Air Blank	0.000	18:32	
Subject Sample	# 2	0.000	18:33
Breath Volume	2.167	18:33	
Air Blank	0.000	18:33	
Air Blank	0.000	18:35	
Subject Sample	# 3	0.000	18:36
Breath Volume	3.664	18:36	
Air Blank	0.000	18:36	
Control Test	0.079	18:37	
Air Blank	0.000	18:37	
Diagnostics Check	OK	18:37	

Data Download Date/Time: 03/30/2017 20:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHOTSCH,RYAN,A
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: SCHOTSCH,RYAN
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 02/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629801
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 22:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:55

SUBJECT'S LAST NAME: WHEELER

DATE/TIME OF TEST: 03/17/2017 23:21:31

Diagnostics Check	OK	23:17	
Air Blank	0.000	23:18	
Control Test	0.079	23:18	
Air Blank	0.000	23:19	
Subject Sample	# 1	23:20	Subject Test Refused
Breath Volume	0.000	23:20	
Air Blank	0.000	23:20	
Control Test	0.078	23:20	
Air Blank	0.000	23:21	
Diagnostics Check	OK	23:21	

Data Download Date/Time: 03/30/2017 20:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPENCER, RAYMOND, L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: SPENCER, RAYMOND
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 02/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629801
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: ANDREWS

DATE/TIME OF TEST: 03/19/2017 01:18:06

Diagnostics Check	OK	01:11	
Air Blank	0.000	01:11	
Control Test	0.080	01:11	
Air Blank	0.000	01:12	
Subject Sample	# 1 0.046	01:14	
Breath Volume	1.468	01:14	
Air Blank	0.000	01:15	
Air Blank	0.000	01:17	
Subject Sample	# 2	01:17	RFI Detect
Breath Volume	0.000	01:17	
Air Blank		01:17	RFI Detect
Air Blank	0.000	01:18	

Data Download Date/Time: 03/30/2017 20:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPENCER, RAYMOND, L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: SPENCER, RAYMOND
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 02/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629801
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: ANDREWS

DATE/TIME OF TEST: 03/19/2017 01:26:49

Diagnostics Check	OK	01:20
Air Blank	0.000	01:20
Control Test	0.080	01:20
Air Blank	0.000	01:21
Subject Sample # 1	0.044	01:22
Breath Volume	1.453	01:22
Air Blank	0.000	01:22
Air Blank	0.000	01:24
Subject Sample # 2	0.045	01:25
Breath Volume	1.472	01:25
Air Blank	0.000	01:25
Control Test	0.080	01:26
Air Blank	0.000	01:26
Diagnostics Check	OK	01:26

Data Download Date/Time: 03/30/2017 20:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VITALE,MICHAEL,L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: VITALE,MICHAEL
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 02/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629801
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 04:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:31

SUBJECT'S LAST NAME: KELLEY

DATE/TIME OF TEST: 03/19/2017 06:05:37

Diagnostics Check	OK	05:54	
Air Blank	0.000	05:54	
Control Test	0.080	05:55	
Air Blank	0.000	05:55	
Subject Sample	# 1	05:58	No Sample Provided
Breath Volume	0.000	05:58	
Air Blank	0.000	05:59	
Air Blank	0.000	06:01	
Subject Sample	# 2	06:04	No Sample Provided
Breath Volume	0.000	06:04	
Air Blank	0.000	06:04	
Control Test	0.080	06:05	
Air Blank	0.000	06:05	
Diagnostics Check	OK	06:05	

Data Download Date/Time: 03/30/2017 20:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALLACE,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: WALLACE,CHRISTOPHER
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 02/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629801
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 10:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:50

SUBJECT'S LAST NAME: RUGGS

DATE/TIME OF TEST: 03/21/2017 12:39:53

Diagnostics Check	OK	12:34
Air Blank	0.000	12:34
Control Test	0.079	12:34
Air Blank	0.000	12:35
Subject Sample # 1	0.000	12:35
Breath Volume	3.503	12:35
Air Blank	0.000	12:36
Air Blank	0.000	12:38
Subject Sample # 2	0.000	12:38
Breath Volume	4.273	12:38
Air Blank	0.000	12:38
Control Test	0.079	12:39
Air Blank	0.000	12:39
Diagnostics Check	OK	12:39

Data Download Date/Time: 03/30/2017 20:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,ASHLEY,E
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: THOMPSON,ASHLEY
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 02/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629801
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 14:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:15

SUBJECT'S LAST NAME: VEAZEY

DATE/TIME OF TEST: 03/21/2017 15:50:40

Diagnostics Check	OK	15:39	
Air Blank	0.000	15:39	
Control Test	0.079	15:40	
Air Blank	0.000	15:40	
Subject Sample	# 1	15:43	No Sample Provided
Breath Volume	0.000	15:43	
Air Blank	0.000	15:44	
Air Blank	0.000	15:46	
Subject Sample	# 2	15:49	No Sample Provided
Breath Volume	0.000	15:49	
Air Blank	0.000	15:49	
Control Test	0.079	15:50	
Air Blank	0.000	15:50	
Diagnostics Check	OK	15:50	

Data Download Date/Time: 03/30/2017 20:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LYNCH, CHAD, J
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: LYNCH, CHAD
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 02/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629801
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 19:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:43

SUBJECT'S LAST NAME: BILLOCK

DATE/TIME OF TEST: 03/22/2017 21:12:13

Diagnostics Check	OK	21:03	
Air Blank	0.000	21:04	
Control Test	0.080	21:04	
Air Blank	0.000	21:05	
Subject Sample # 1	0.159	21:06	
Breath Volume	2.238	21:06	
Air Blank	0.000	21:07	
Air Blank	0.000	21:09	
Subject Sample # 2	0.154	21:10	
Breath Volume	1.324	21:10	
Air Blank	0.000	21:11	
Control Test	0.079	21:11	
Air Blank		21:11	RFI Detect
Air Blank	0.000	21:12	

Data Download Date/Time: 03/30/2017 20:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLEMENTS,WILLIAM,J
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: CLEMENTS,WILLIAM
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 02/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629801
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 16:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:52

SUBJECT'S LAST NAME: MULLIN

DATE/TIME OF TEST: 03/23/2017 17:17:55

Diagnostics Check	OK	17:14	
Air Blank	0.000	17:14	
Control Test	0.080	17:15	
Air Blank	0.000	17:15	
Subject Sample	# 1	17:16	Subject Test Refused
Breath Volume	0.000	17:16	
Air Blank	0.000	17:17	
Control Test	0.080	17:17	
Air Blank	0.000	17:17	
Diagnostics Check	OK	17:17	

Data Download Date/Time: 03/30/2017 20:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUICK,STEPHEN,W
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: QUICK,STEPHEN
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 02/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629801
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 22:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:19

SUBJECT'S LAST NAME: WATZ

DATE/TIME OF TEST: 03/24/2017 23:47:11

Diagnostics Check	OK	23:40
Air Blank	0.000	23:41
Control Test	0.079	23:41
Air Blank	0.000	23:41
Subject Sample # 1	0.089	23:42
Breath Volume	3.484	23:42
Air Blank	0.000	23:43
Air Blank	0.000	23:45
Subject Sample # 2	0.091	23:45
Breath Volume	3.171	23:45
Air Blank	0.000	23:46
Control Test	0.079	23:46
Air Blank	0.000	23:47
Diagnostics Check	OK	23:47

Data Download Date/Time: 03/30/2017 20:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,JOHN,F
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: YOUNG,JOHN
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 02/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629801
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 23:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:39

SUBJECT'S LAST NAME: ROBBINS

DATE/TIME OF TEST: 03/27/2017 00:06:01

Diagnostics Check	OK	23:59
Air Blank	0.000	00:00
Control Test	0.079	00:00
Air Blank	0.000	00:00
Subject Sample # 1	0.179	00:01
Breath Volume	3.472	00:01
Air Blank	0.000	00:02
Air Blank	0.000	00:04
Subject Sample # 2	0.187	00:04
Breath Volume	3.765	00:04
Air Blank	0.000	00:05
Control Test	0.078	00:05
Air Blank	0.000	00:05
Diagnostics Check	OK	00:06

Data Download Date/Time: 03/30/2017 20:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACKSON,VICTOR,A
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: JACKSON,VICTOR
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 02/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629801
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 16:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:21

SUBJECT'S LAST NAME: BRYSON

DATE/TIME OF TEST: 03/27/2017 17:52:04

Diagnostics Check	OK	17:44
Air Blank	0.000	17:44
Control Test	0.078	17:44
Air Blank	0.000	17:45
Subject Sample # 1	0.092	17:47
Breath Volume	1.992	17:47
Air Blank	0.000	17:47
Air Blank	0.000	17:49
Subject Sample # 2	0.084	17:50
Breath Volume	2.234	17:50
Air Blank	0.000	17:51
Control Test	0.078	17:51
Air Blank	0.000	17:51
Diagnostics Check	OK	17:52

Data Download Date/Time: 03/30/2017 20:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHAPMAN,MAVERICK,E
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: CHAPMAN,MAVERICK
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 02/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629801
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 05:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:35

SUBJECT'S LAST NAME: MCCOY

DATE/TIME OF TEST: 03/28/2017 06:02:15

Diagnostics Check	OK	05:55
Air Blank	0.000	05:56
Control Test	0.079	05:56
Air Blank	0.000	05:57
Subject Sample # 1	0.000	05:57
Breath Volume	1.550	05:57
Air Blank	0.000	05:58
Air Blank	0.000	06:00
Subject Sample # 2	0.000	06:00
Breath Volume	1.605	06:00
Air Blank	0.000	06:01
Control Test	0.079	06:01
Air Blank	0.000	06:02
Diagnostics Check	OK	06:02

Data Download Date/Time: 03/30/2017 20:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BENNETT,EDGAR,A
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: BENNETT,EDGAR
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 02/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629801
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 02:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: SENAU

DATE/TIME OF TEST: 03/29/2017 03:27:32

Diagnostics Check	OK	03:24	
Air Blank	0.000	03:24	
Control Test	0.079	03:24	
Air Blank	0.000	03:25	
Subject Sample	# 1	03:26	Subject Test Refused
Breath Volume	0.000	03:26	
Air Blank	0.000	03:26	
Control Test	0.079	03:26	
Air Blank	0.000	03:27	
Diagnostics Check	OK	03:27	

Data Download Date/Time: 03/30/2017 20:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DANIELS,DAVID,C
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: DANIELS,DAVID
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 02/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629801
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 03:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:50

SUBJECT'S LAST NAME: GRACE

DATE/TIME OF TEST: 03/29/2017 05:17:00

Diagnostics Check	OK	05:11
Air Blank	0.000	05:11
Control Test	0.079	05:11
Air Blank	0.000	05:12
Subject Sample # 1	0.000	05:12
Breath Volume	1.929	05:12
Air Blank	0.000	05:13
Air Blank	0.000	05:15
Subject Sample # 2	0.000	05:15
Breath Volume	1.519	05:15
Air Blank	0.000	05:16
Control Test	0.078	05:16
Air Blank	0.000	05:16
Diagnostics Check	OK	05:17

Data Download Date/Time: 03/30/2017 20:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPBELL,HENDESON,S
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: CAMPBELL,HENDERSON
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 02/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629801
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 20:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:33

SUBJECT'S LAST NAME: MACKIE

DATE/TIME OF TEST: 03/29/2017 22:02:09

Diagnostics Check	OK	21:55	
Air Blank	0.000	21:56	
Control Test	0.079	21:56	
Air Blank	0.000	21:56	
Subject Sample	# 1 0.216	21:57	
Breath Volume	3.578	21:57	
Air Blank	0.000	21:58	
Air Blank	0.000	22:00	
Subject Sample	# 2 0.209	22:01	
Breath Volume	2.875	22:01	
Air Blank		22:01	RFI Detect
Air Blank	0.000	22:02	

Data Download Date/Time: 03/30/2017 20:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOINER,JEREMY,R
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: BODNOVITS,JASON
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 03/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629801
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 23:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: JOINER

DATE/TIME OF TEST: 03/30/2017 23:46:26

Diagnostics Check	OK	23:45	
Air Blank		23:45	RFI Detect
Air Blank	0.000	23:46	

Data Download Date/Time: 04/30/2017 22:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZICARI,GAETANO,S
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: BODNOVITS,JASON
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 03/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629801
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 23:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: JOINER

DATE/TIME OF TEST: 03/30/2017 23:56:31

Diagnostics Check	OK	23:48
Air Blank	0.000	23:49
Control Test	0.079	23:49
Air Blank	0.000	23:49
Subject Sample # 1	0.257	23:50
Breath Volume	1.132	23:50
Air Blank	0.000	23:51
Air Blank	0.000	23:53
Subject Sample # 2	0.239	23:54
Breath Volume	1.660	23:54
Air Blank	0.000	23:55
Control Test	0.079	23:55
Air Blank	0.000	23:56
Diagnostics Check	OK	23:56

Data Download Date/Time: 04/30/2017 22:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRIBB,BRIAN,S
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: CRIBB,BRIAN
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 03/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629801
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 20:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:12

SUBJECT'S LAST NAME: MULLICAN

DATE/TIME OF TEST: 03/31/2017 21:39:11

Diagnostics Check	OK	21:32
Air Blank	0.000	21:33
Control Test	0.080	21:33
Air Blank	0.000	21:33
Subject Sample # 1	0.000	21:34
Breath Volume	1.441	21:34
Air Blank	0.000	21:35
Air Blank	0.000	21:37
Subject Sample # 2	0.000	21:37
Breath Volume	1.691	21:37
Air Blank	0.000	21:38
Control Test	0.079	21:38
Air Blank	0.000	21:39
Diagnostics Check	OK	21:39

Data Download Date/Time: 04/30/2017 22:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001305

INSTRUMENT REGISTERED TO: MARIANNA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOLDEN,ANTHONY,E
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: GOLDEN,ANTHONY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 02/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1650824
DRY GAS EXPIRY DATE: 03/13/2017
VIOLATION CODE: DUI
ARREST TIME: 19:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:30

SUBJECT'S LAST NAME: REYNOLDS

DATE/TIME OF TEST: 03/01/2017 20:22:38

Diagnostics Check	OK	20:11	
Air Blank	0.000	20:11	
Control Test	0.080	20:11	
Air Blank	0.000	20:12	
Subject Sample	# 1 0.157	20:15	Volume Not Met
Breath Volume	1.105	20:15	
Air Blank	0.000	20:16	
Air Blank	0.000	20:17	
Subject Sample	# 2 0.153	20:21	Volume Not Met
Breath Volume	0.597	20:21	
Air Blank	0.000	20:21	
Control Test	0.079	20:22	
Air Blank	0.000	20:22	
Diagnostics Check	OK	20:22	

Data Download Date/Time: 03/27/2017 16:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001305

INSTRUMENT REGISTERED TO: MARIANNA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRIFFITH,DALTON,M
BREATH TEST OPERATOR AGENCY: Marianna PD
ARRESTING OFFICER: GRIFFITH,DALTON
ARRESTING OFFICER AGENCY: Marianna PD
LAST AGENCY INSPECTION: 02/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 01:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: HARDING

DATE/TIME OF TEST: 03/15/2017 02:21:07

Diagnostics Check	OK	02:06	
Air Blank	0.000	02:07	
Control Test	0.081	02:07	
Air Blank	0.000	02:08	
Subject Sample	# 1	02:11	No Sample Provided
Breath Volume	0.000	02:11	
Air Blank	0.000	02:11	
Air Blank	0.000	02:13	
Subject Sample	# 2	0.000	02:16
Breath Volume	2.484	02:16	
Air Blank	0.000	02:16	
Air Blank	0.000	02:18	
Subject Sample	# 3	0.000	02:19
Breath Volume	2.027	02:19	
Air Blank	0.000	02:20	
Control Test	0.081	02:20	
Air Blank	0.000	02:21	
Diagnostics Check	OK	02:21	

Data Download Date/Time: 03/27/2017 16:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001305

INSTRUMENT REGISTERED TO: MARIANNA PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: RIGGINS,MYRON,D

BREATH TEST OPERATOR AGENCY: FHP Troop A

ARRESTING OFFICER: MEALOR,THOMAS

ARRESTING OFFICER AGENCY: FHP Troop A

LAST AGENCY INSPECTION: 02/21/2017

Date or Time Changed: No

DRY GAS STANDARD LOT #: 666491

DRY GAS EXPIRY DATE: 12/30/2019

VIOLATION CODE: DUI

ARREST TIME: 12:39

GENDER: Male

OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: RIGGINS

DATE/TIME OF TEST: 03/16/2017 02:24:00

Diagnostics Check	OK	02:16
Air Blank	0.000	02:17
Control Test	0.082	02:17
Air Blank	0.000	02:18
Subject Sample # 1	0.135	02:19
Breath Volume	1.339	02:19
Air Blank	0.000	02:19
Air Blank	0.000	02:21
Subject Sample # 2	0.136	02:22
Breath Volume	1.539	02:22
Air Blank	0.000	02:23
Control Test	0.081	02:23
Air Blank	0.000	02:23
Diagnostics Check	OK	02:24

Data Download Date/Time: 03/27/2017 16:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001305

INSTRUMENT REGISTERED TO: MARIANNA PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BOWMAN,JOANNA,N

BREATH TEST OPERATOR AGENCY: FHP Troop A

ARRESTING OFFICER: BOWMAN,JOANNA

ARRESTING OFFICER AGENCY: FHP Troop A

LAST AGENCY INSPECTION: 02/21/2017

Date or Time Changed: No

DRY GAS STANDARD LOT #: 666491

DRY GAS EXPIRY DATE: 12/30/2019

VIOLATION CODE: DUI

ARREST TIME: 07:36

GENDER: Female

OBSERVATION PERIOD BEGAN: 07:36

SUBJECT'S LAST NAME: BAREFOOT

DATE/TIME OF TEST: 03/19/2017 09:25:44

Diagnostics Check	OK	09:18
Air Blank	0.000	09:18
Control Test	0.080	09:19
Air Blank	0.000	09:19
Subject Sample # 1	0.077	09:21
Breath Volume	1.429	09:21
Air Blank	0.000	09:21
Air Blank	0.000	09:23
Subject Sample # 2	0.075	09:24
Breath Volume	1.695	09:24
Air Blank	0.000	09:24
Control Test	0.080	09:25
Air Blank	0.000	09:25
Diagnostics Check	OK	09:25

Data Download Date/Time: 03/27/2017 16:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001305

INSTRUMENT REGISTERED TO: MARIANNA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOLDEN,ANTHONY,E
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: GOLDEN,ANTHONY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 03/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 15:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:36

SUBJECT'S LAST NAME: HUDGINS

DATE/TIME OF TEST: 03/26/2017 17:08:28

Diagnostics Check	OK	17:01
Air Blank	0.000	17:02
Control Test	0.081	17:02
Air Blank	0.000	17:03
Subject Sample # 1	0.123	17:03
Breath Volume	1.976	17:03
Air Blank	0.000	17:04
Air Blank	0.000	17:05
Subject Sample # 2	0.124	17:06
Breath Volume	2.371	17:06
Air Blank	0.000	17:07
Control Test	0.081	17:07
Air Blank	0.000	17:08
Diagnostics Check	OK	17:08

Data Download Date/Time: 03/27/2017 16:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001305

INSTRUMENT REGISTERED TO: MARIANNA PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HARSEY, WESLEY, H

BREATH TEST OPERATOR AGENCY: FHP Troop A

ARRESTING OFFICER: HARSEY, WESLEY

ARRESTING OFFICER AGENCY: FHP Troop A

LAST AGENCY INSPECTION: 03/24/2017

Date or Time Changed: No

DRY GAS STANDARD LOT #: 666491

DRY GAS EXPIRY DATE: 12/30/2019

VIOLATION CODE: DUI

ARREST TIME: 13:44

GENDER: Male

OBSERVATION PERIOD BEGAN: 14:30

SUBJECT'S LAST NAME: SHERMAN

DATE/TIME OF TEST: 03/27/2017 15:06:21

Diagnostics Check	OK	15:03	
Air Blank	0.000	15:03	
Control Test	0.080	15:03	
Air Blank	0.000	15:04	
Subject Sample	# 1	15:04	Subject Test Refused
Breath Volume	0.000	15:04	
Air Blank	0.000	15:05	
Control Test	0.081	15:05	
Air Blank	0.000	15:06	
Diagnostics Check	OK	15:06	

Data Download Date/Time: 03/27/2017 16:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001306

INSTRUMENT REGISTERED TO: MEXICO BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOUSE,STEVEN,D
BREATH TEST OPERATOR AGENCY: Mexico Beach PD
ARRESTING OFFICER: FOSTER,ALEX
ARRESTING OFFICER AGENCY: Mexico Beach PD
LAST AGENCY INSPECTION: 03/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1837677
DRY GAS EXPIRY DATE: 04/02/2018
VIOLATION CODE: DUI
ARREST TIME: 16:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:17

SUBJECT'S LAST NAME: HARRELL

DATE/TIME OF TEST: 03/07/2017 17:56:26

Diagnostics Check	OK	17:49
Air Blank	0.000	17:50
Control Test	0.078	17:50
Air Blank	0.000	17:51
Subject Sample # 1	0.266	17:51
Breath Volume	2.457	17:51
Air Blank	0.000	17:52
Air Blank	0.000	17:54
Subject Sample # 2	0.261	17:54
Breath Volume	2.714	17:54
Air Blank	0.000	17:55
Control Test	0.079	17:55
Air Blank	0.000	17:56
Diagnostics Check	OK	17:56

Data Download Date/Time: 04/03/2017 02:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001306

INSTRUMENT REGISTERED TO: MEXICO BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCSWINEY,RICHARD,A
BREATH TEST OPERATOR AGENCY: Mexico Beach PD
ARRESTING OFFICER: MCSWINEY,RICHARD
ARRESTING OFFICER AGENCY: Mexico Beach PD
LAST AGENCY INSPECTION: 03/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1837677
DRY GAS EXPIRY DATE: 04/02/2018
VIOLATION CODE: DUI
ARREST TIME: 22:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:25

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 03/14/2017 23:00:32

Diagnostics Check	OK	22:54
Air Blank	0.000	22:55
Control Test	0.080	22:55
Air Blank	0.000	22:55
Subject Sample # 1	0.224	22:56
Breath Volume	3.703	22:56
Air Blank	0.000	22:56
Air Blank	0.000	22:58
Subject Sample # 2	0.228	22:59
Breath Volume	3.496	22:59
Air Blank	0.000	22:59
Control Test	0.080	22:59
Air Blank	0.000	23:00
Diagnostics Check	OK	23:00

Data Download Date/Time: 04/03/2017 02:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001307

INSTRUMENT REGISTERED TO: MADISON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRELL, WILLIAM, E
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: HARRELL, WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 03/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: ODOM

DATE/TIME OF TEST: 03/15/2017 00:08:21

Diagnostics Check	OK	00:05	
Air Blank	0.000	00:05	
Control Test	0.078	00:05	
Air Blank	0.000	00:06	
Subject Sample	# 1	00:06	Subject Test Refused
Breath Volume	0.000	00:06	
Air Blank	0.000	00:07	
Control Test	0.079	00:07	
Air Blank	0.000	00:08	
Diagnostics Check	OK	00:08	

Data Download Date/Time: 04/10/2017 10:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOWMAN,JOANNA,N
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BOWMAN,JOANNA
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 02/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG628401
DRY GAS EXPIRY DATE: 10/10/2018
VIOLATION CODE: DUI
ARREST TIME: 03:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:17

SUBJECT'S LAST NAME: BOTTORF

DATE/TIME OF TEST: 03/04/2017 03:59:40

Diagnostics Check	OK	03:54	
Air Blank	0.000	03:55	
Control Test	0.079	03:55	
Air Blank	0.000	03:56	
Subject Sample	# 1	03:59	Sequence Aborted
Breath Volume	0.910	03:59	
Air Blank	0.000	03:59	

Data Download Date/Time: 04/03/2017 18:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUBEL,NEIL,F
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: RUBEL,NEIL
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 02/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG628401
DRY GAS EXPIRY DATE: 10/10/2018
VIOLATION CODE: DUI
ARREST TIME: 02:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:46

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 03/09/2017 03:09:30

Diagnostics Check	OK	03:06	
Air Blank	0.000	03:06	
Control Test	0.079	03:07	
Air Blank	0.000	03:07	
Subject Sample	# 1	03:08	Subject Test Refused
Breath Volume	0.000	03:08	
Air Blank	0.000	03:08	
Control Test	0.078	03:08	
Air Blank	0.000	03:09	
Diagnostics Check	OK	03:09	

Data Download Date/Time: 04/03/2017 18:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SOMERS,TAMMY,L
BREATH TEST OPERATOR AGENCY: Parker PD
ARRESTING OFFICER: SUMMERS,EDWARD
ARRESTING OFFICER AGENCY: Parker PD
LAST AGENCY INSPECTION: 02/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG628401
DRY GAS EXPIRY DATE: 10/10/2018
VIOLATION CODE: DUI
ARREST TIME: 19:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:40

SUBJECT'S LAST NAME: SOMERS

DATE/TIME OF TEST: 03/11/2017 21:14:16

Diagnostics Check	OK	21:00	
Air Blank	0.000	21:01	
Control Test	0.079	21:01	
Air Blank	0.000	21:01	
Subject Sample	# 1 0.119	21:02	
Breath Volume	1.101	21:02	
Air Blank	0.000	21:03	
Air Blank	0.000	21:05	
Subject Sample	# 2 0.113	21:08	Volume Not Met
Breath Volume	0.175	21:08	
Air Blank	0.000	21:08	
Air Blank	0.000	21:10	
Subject Sample	# 3 0.117	21:12	
Breath Volume	1.253	21:12	
Air Blank	0.000	21:13	
Control Test	0.078	21:13	
Air Blank	0.000	21:14	
Diagnostics Check	OK	21:14	

Data Download Date/Time: 04/03/2017 18:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAMMONS,CHRISTOPHER,R
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: HOUSE,S
ARRESTING OFFICER AGENCY: Mexico Beach PD
LAST AGENCY INSPECTION: 02/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG628401
DRY GAS EXPIRY DATE: 10/10/2018
VIOLATION CODE: DUI
ARREST TIME: 01:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: HOGAN

DATE/TIME OF TEST: 03/12/2017 04:42:28

Diagnostics Check	OK	04:39	
Air Blank	0.000	04:40	
Control Test	0.079	04:40	
Air Blank	0.000	04:40	
Subject Sample	# 1	04:41	Subject Test Refused
Breath Volume	0.000	04:41	
Air Blank	0.000	04:41	
Control Test	0.079	04:41	
Air Blank	0.000	04:42	
Diagnostics Check	OK	04:42	

Data Download Date/Time: 04/03/2017 18:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROSIUS,STANLEY,W
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: PARKER,C
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 02/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG628401
DRY GAS EXPIRY DATE: 10/10/2018
VIOLATION CODE: DUI
ARREST TIME: 18:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:18

SUBJECT'S LAST NAME: YOUNGBLOOD-GREEN

DATE/TIME OF TEST: 03/15/2017 19:42:06

Diagnostics Check	OK	19:39	
Air Blank	0.000	19:39	
Control Test	0.079	19:40	
Air Blank	0.000	19:40	
Subject Sample	# 1	19:40	Subject Test Refused
Breath Volume	0.000	19:40	
Air Blank	0.000	19:41	
Control Test	0.079	19:41	
Air Blank	0.000	19:42	
Diagnostics Check	OK	19:42	

Data Download Date/Time: 04/03/2017 18:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOHENY,K,
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: FIRKS,M
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 02/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG628401
DRY GAS EXPIRY DATE: 10/10/2018
VIOLATION CODE: DUI
ARREST TIME: 21:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: REID

DATE/TIME OF TEST: 03/16/2017 22:11:45

Diagnostics Check	OK	22:05
Air Blank	0.000	22:05
Control Test	0.078	22:06
Air Blank	0.000	22:06
Subject Sample # 1	0.364	22:07
Breath Volume	2.960	22:07
Air Blank	0.000	22:07
Air Blank	0.000	22:09
Subject Sample # 2	0.353	22:10
Breath Volume	2.531	22:10
Air Blank	0.000	22:10
Control Test	0.077	22:11
Air Blank	0.000	22:11
Diagnostics Check	OK	22:11

Data Download Date/Time: 04/03/2017 18:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRIS,DAVID,L
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: HARRIS,DAVID
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 02/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG628401
DRY GAS EXPIRY DATE: 10/10/2018
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:28

SUBJECT'S LAST NAME: DUDLEY

DATE/TIME OF TEST: 03/17/2017 01:05:17

Diagnostics Check	OK	00:57	
Air Blank	0.000	00:58	
Control Test	0.080	00:58	
Air Blank	0.000	00:58	
Subject Sample	# 1 0.140	01:02	
Breath Volume	1.152	01:02	
Air Blank	0.000	01:02	
Air Blank	0.000	01:04	
		01:04	RFI Detect
Air Blank		01:04	RFI Detect
Air Blank	0.000	01:05	

Data Download Date/Time: 04/03/2017 18:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRIS,DAVID,L
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: HARRIS,DAVID
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 02/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG628401
DRY GAS EXPIRY DATE: 10/10/2018
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: DUDLEY

DATE/TIME OF TEST: 03/17/2017 01:14:31

Diagnostics Check	OK	01:07
Air Blank	0.000	01:07
Control Test	0.079	01:07
Air Blank	0.000	01:08
Subject Sample # 1	0.127	01:09
Breath Volume	1.628	01:09
Air Blank	0.000	01:10
Air Blank	0.000	01:12
Subject Sample # 2	0.128	01:13
Breath Volume	1.378	01:13
Air Blank	0.000	01:13
Control Test	0.078	01:13
Air Blank	0.000	01:14
Diagnostics Check	OK	01:14

Data Download Date/Time: 04/03/2017 18:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROSIUS,STANLEY,W
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: THOMPSON,J
ARRESTING OFFICER AGENCY: Lynn Haven PD
LAST AGENCY INSPECTION: 02/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG628401
DRY GAS EXPIRY DATE: 10/10/2018
VIOLATION CODE: DUI
ARREST TIME: 00:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:18

SUBJECT'S LAST NAME: RUSHING

DATE/TIME OF TEST: 03/21/2017 01:47:49

Diagnostics Check	OK	01:40
Air Blank	0.000	01:41
Control Test	0.079	01:41
Air Blank	0.000	01:42
Subject Sample # 1	0.106	01:43
Breath Volume	2.781	01:43
Air Blank	0.000	01:43
Air Blank	0.000	01:45
Subject Sample # 2	0.098	01:46
Breath Volume	1.667	01:46
Air Blank	0.000	01:46
Control Test	0.078	01:47
Air Blank	0.000	01:47
Diagnostics Check	OK	01:47

Data Download Date/Time: 04/03/2017 18:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROSIUS,STANLEY,W
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: GOBER,J
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 02/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG628401
DRY GAS EXPIRY DATE: 10/10/2018
VIOLATION CODE: DUI
ARREST TIME: 02:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:48

SUBJECT'S LAST NAME: HOWARD

DATE/TIME OF TEST: 03/26/2017 03:14:04

Diagnostics Check	OK	03:11	
Air Blank	0.000	03:11	
Control Test	0.079	03:12	
Air Blank	0.000	03:12	
Subject Sample	# 1	03:12	Subject Test Refused
Breath Volume	0.000	03:12	
Air Blank	0.000	03:13	
Control Test	0.079	03:13	
Air Blank	0.000	03:13	
Diagnostics Check	OK	03:14	

Data Download Date/Time: 04/03/2017 18:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROSIUS,STANLEY,W
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: WARREN,P
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 02/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG628401
DRY GAS EXPIRY DATE: 10/10/2018
VIOLATION CODE: DUI
ARREST TIME: 01:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:28

SUBJECT'S LAST NAME: NEWTON

DATE/TIME OF TEST: 03/27/2017 02:54:56

Diagnostics Check	OK	02:48
Air Blank	0.000	02:48
Control Test	0.078	02:49
Air Blank	0.000	02:49
Subject Sample # 1	0.209	02:50
Breath Volume	1.230	02:50
Air Blank	0.000	02:51
Air Blank	0.000	02:52
Subject Sample # 2	0.209	02:53
Breath Volume	2.222	02:53
Air Blank	0.000	02:54
Control Test	0.077	02:54
Air Blank	0.000	02:54
Diagnostics Check	OK	02:54

Data Download Date/Time: 04/03/2017 18:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POTTER,MONTREZ,R
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: POTTER,MONTREZ
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 03/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG628401
DRY GAS EXPIRY DATE: 10/10/2018
VIOLATION CODE: DUI
ARREST TIME: 05:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:27

SUBJECT'S LAST NAME: HARDY

DATE/TIME OF TEST: 03/30/2017 07:34:37

Diagnostics Check	OK	07:23	
Air Blank	0.000	07:23	
Control Test	0.078	07:23	
Air Blank	0.000	07:24	
Subject Sample	# 1 0.242	07:27	Volume Not Met
Breath Volume	0.464	07:27	
Air Blank	0.000	07:28	
Air Blank	0.000	07:29	
Subject Sample	# 2	07:32	Subject Test Refused
Breath Volume	0.652	07:32	
Air Blank	0.000	07:33	
Control Test	0.076	07:34	
Air Blank	0.000	07:34	
Diagnostics Check	OK	07:34	

Data Download Date/Time: 04/03/2017 18:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROSIUS,STANLEY,W
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: BYRD,B
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 03/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG628401
DRY GAS EXPIRY DATE: 10/10/2018
VIOLATION CODE: DUI
ARREST TIME: 00:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: DENNEY

DATE/TIME OF TEST: 03/31/2017 01:25:04

Diagnostics Check	OK	01:11	
Air Blank	0.000	01:11	
Control Test	0.078	01:11	
Air Blank	0.000	01:12	
Subject Sample	# 1 0.153	01:15	Volume Not Met
Breath Volume	0.472	01:15	
Air Blank	0.000	01:16	
Air Blank	0.000	01:18	
Subject Sample	# 2 0.198	01:20	
Breath Volume	2.695	01:20	
Air Blank	0.000	01:21	
Air Blank	0.000	01:23	
Subject Sample	# 3 0.193	01:23	
Breath Volume	1.535	01:23	
Air Blank	0.000	01:24	
Control Test	0.076	01:24	
Air Blank	0.000	01:24	
Diagnostics Check	OK	01:25	

Data Download Date/Time: 04/03/2017 18:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001310

INSTRUMENT REGISTERED TO: NICEVILLE PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: CRAGHEAD,JEREMY,S

BREATH TEST OPERATOR AGENCY: Valparaiso PD

ARRESTING OFFICER: CRAGHEAD,JEREMY

ARRESTING OFFICER AGENCY: Valparaiso PD

LAST AGENCY INSPECTION: 02/28/2017

Date or Time Changed: No

DRY GAS STANDARD LOT #: 507129

DRY GAS EXPIRY DATE: 08/04/2019

VIOLATION CODE: DUI

ARREST TIME: 23:15

GENDER: Male

OBSERVATION PERIOD BEGAN: 23:26

SUBJECT'S LAST NAME: FINNERTY

DATE/TIME OF TEST: 03/07/2017 23:53:10

Diagnostics Check	OK	23:50	
Air Blank	0.000	23:50	
Control Test	0.083	23:51	
Air Blank	0.000	23:51	
Subject Sample	# 1	23:51	Subject Test Refused
Breath Volume	0.000	23:51	
Air Blank	0.000	23:52	
Control Test	0.084	23:52	
Air Blank	0.000	23:53	
Diagnostics Check	OK	23:53	

Data Download Date/Time: 04/06/2017 03:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001310

INSTRUMENT REGISTERED TO: NICEVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEWIS,SHAWN,W
BREATH TEST OPERATOR AGENCY: Niceville PD
ARRESTING OFFICER: MACBETH,MATTHEW
ARRESTING OFFICER AGENCY: Niceville PD
LAST AGENCY INSPECTION: 02/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: DUI
ARREST TIME: 20:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:26

SUBJECT'S LAST NAME: PRESLEY

DATE/TIME OF TEST: 03/10/2017 02:03:51

Diagnostics Check	OK	01:52	
Air Blank	0.000	01:52	
Control Test	0.082	01:53	
Air Blank	0.000	01:53	
Subject Sample	# 1 0.143	01:56	Volume Not Met
Breath Volume	0.175	01:56	
Air Blank	0.000	01:57	
Air Blank	0.000	01:59	
Subject Sample	# 2 0.142	02:02	Volume Not Met
Breath Volume	0.269	02:02	
Air Blank	0.000	02:02	
Control Test	0.080	02:03	
Air Blank	0.000	02:03	
Diagnostics Check	OK	02:03	

Data Download Date/Time: 04/06/2017 03:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001310

INSTRUMENT REGISTERED TO: NICEVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SULLIVAN,DANNY,
BREATH TEST OPERATOR AGENCY: Eglin AFB PD
ARRESTING OFFICER: FOY,CALEB
ARRESTING OFFICER AGENCY: Eglin AFB PD
LAST AGENCY INSPECTION: 02/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: DUI
ARREST TIME: 01:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:01

SUBJECT'S LAST NAME: HAMILTON

DATE/TIME OF TEST: 03/18/2017 03:44:52

Diagnostics Check	OK	03:41	
Air Blank	0.000	03:42	
Control Test	0.083	03:42	
Air Blank	0.000	03:43	
Subject Sample	# 1	03:43	Subject Test Refused
Breath Volume	0.000	03:43	
Air Blank	0.000	03:43	
Control Test	0.083	03:44	
Air Blank	0.000	03:44	
Diagnostics Check	OK	03:44	

Data Download Date/Time: 04/06/2017 03:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001310

INSTRUMENT REGISTERED TO: NICEVILLE PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LONGO,SARAH,J

BREATH TEST OPERATOR AGENCY: Eglin AFB PD

ARRESTING OFFICER: FOY,CALEB

ARRESTING OFFICER AGENCY: Eglin AFB PD

LAST AGENCY INSPECTION: 02/28/2017

Date or Time Changed: Yes

DRY GAS STANDARD LOT #: 507129

DRY GAS EXPIRY DATE: 08/04/2019

VIOLATION CODE: DUI

ARREST TIME: 12:03

GENDER: Male

OBSERVATION PERIOD BEGAN: 01:57

SUBJECT'S LAST NAME: ARCHER

DATE/TIME OF TEST: 03/25/2017 02:33:29

Diagnostics Check	OK	02:20	
Air Blank	0.000	02:21	
Control Test	0.082	02:21	
Air Blank	0.000	02:21	
Subject Sample	# 1 0.140	02:24	Volume Not Met
Breath Volume	0.472	02:24	
Air Blank	0.000	02:25	
Air Blank	0.000	02:27	
Subject Sample	# 2 0.146	02:28	
Breath Volume	1.832	02:28	
Air Blank	0.000	02:29	
Air Blank	0.000	02:30	
Subject Sample	# 3 0.142	02:31	
Breath Volume	1.792	02:31	
Air Blank	0.000	02:32	
Control Test	0.080	02:32	
Air Blank	0.000	02:33	
Diagnostics Check	OK	02:33	

Data Download Date/Time: 04/06/2017 03:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KEMP,MARK,A
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: KEMP,MARK
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 02/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 14:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:38

SUBJECT'S LAST NAME: WYLLIE

DATE/TIME OF TEST: 03/13/2017 16:07:19

Diagnostics Check	OK	16:00
Air Blank	0.000	16:01
Control Test	0.081	16:01
Air Blank	0.000	16:02
Subject Sample # 1	0.320	16:02
Breath Volume	3.242	16:02
Air Blank	0.000	16:03
Air Blank	0.000	16:05
Subject Sample # 2	0.323	16:05
Breath Volume	3.929	16:05
Air Blank	0.000	16:06
Control Test	0.079	16:06
Air Blank	0.000	16:07
Diagnostics Check	OK	16:07

Data Download Date/Time: 03/29/2017 15:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRENCH,JASON,K
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: FRENCH,JASON
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 02/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 09:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:50

SUBJECT'S LAST NAME: OHAGEN

DATE/TIME OF TEST: 03/15/2017 10:18:13

Diagnostics Check	OK	10:11
Air Blank	0.000	10:11
Control Test	0.081	10:12
Air Blank	0.000	10:12
Subject Sample # 1	0.196	10:13
Breath Volume	2.660	10:13
Air Blank	0.000	10:14
Air Blank	0.000	10:15
Subject Sample # 2	0.182	10:16
Breath Volume	1.316	10:16
Air Blank	0.000	10:17
Control Test	0.080	10:17
Air Blank	0.000	10:18
Diagnostics Check	OK	10:18

Data Download Date/Time: 03/29/2017 15:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANDERS,DAVID,O
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: SANDERS,DAVID
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 02/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163937
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 22:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: RICE

DATE/TIME OF TEST: 03/17/2017 00:14:55

Diagnostics Check	OK	00:11	
Air Blank	0.000	00:12	
Control Test	0.081	00:12	
Air Blank	0.000	00:13	
Subject Sample	# 1	00:13	Subject Test Refused
Breath Volume	0.000	00:13	
Air Blank	0.000	00:14	
Control Test	0.081	00:14	
Air Blank	0.000	00:14	
Diagnostics Check	OK	00:14	

Data Download Date/Time: 03/29/2017 15:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EARLEY,CHAD,D
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: EARLEY,CHAD
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 02/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163937
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 06:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:30

SUBJECT'S LAST NAME: FIELDS

DATE/TIME OF TEST: 03/17/2017 07:03:48

Diagnostics Check	OK	06:58	
Air Blank	0.000	06:59	
Control Test	0.081	06:59	
Air Blank	0.000	06:59	
Subject Sample	# 1 0.050	07:00	
Breath Volume	1.765	07:00	
Air Blank	0.000	07:01	
Air Blank		07:03	Ambient Fail
Air Blank	0.000	07:03	

Data Download Date/Time: 03/29/2017 15:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EARLEY,CHAD,D
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: EARLEY,CHAD
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 02/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163937
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 06:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:30

SUBJECT'S LAST NAME: FIELDS

DATE/TIME OF TEST: 03/17/2017 07:12:45

Diagnostics Check	OK	07:06
Air Blank	0.000	07:06
Control Test	0.080	07:07
Air Blank	0.000	07:07
Subject Sample # 1	0.051	07:08
Breath Volume	3.644	07:08
Air Blank	0.000	07:08
Air Blank	0.000	07:10
Subject Sample # 2	0.050	07:11
Breath Volume	3.675	07:11
Air Blank	0.000	07:11
Control Test	0.079	07:12
Air Blank	0.000	07:12
Diagnostics Check	OK	07:12

Data Download Date/Time: 03/29/2017 15:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITHSON,CYNTHIA,A
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: GUERRA,MARK
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 02/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163937
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 13:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:00

SUBJECT'S LAST NAME: SMITHSON

DATE/TIME OF TEST: 03/22/2017 14:30:10

Diagnostics Check	OK	14:27	
Air Blank	0.000	14:28	
Control Test	0.080	14:28	
Air Blank	0.000	14:29	
Subject Sample	# 1	14:29	Sequence Aborted
Breath Volume	0.000	14:29	
Air Blank	0.000	14:30	

Data Download Date/Time: 03/29/2017 15:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLISON,STEVEN,L
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: ORTEGA,JUAN
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 02/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 00:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:04

SUBJECT'S LAST NAME: DOYLE

DATE/TIME OF TEST: 03/23/2017 02:31:38

Diagnostics Check	OK	02:26	
Air Blank	0.000	02:26	
Control Test	0.082	02:27	
Air Blank	0.000	02:27	
Subject Sample	# 1	02:30	Subject Test Refused
Breath Volume	0.000	02:30	
Air Blank	0.000	02:30	
Control Test	0.081	02:31	
Air Blank	0.000	02:31	
Diagnostics Check	OK	02:31	

Data Download Date/Time: 03/29/2017 15:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURGESS,MASON,W
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BURGESS,MASON
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 02/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: DANIEL

DATE/TIME OF TEST: 03/24/2017 02:30:21

Diagnostics Check	OK	02:23
Air Blank	0.000	02:23
Control Test	0.080	02:24
Air Blank	0.000	02:24
Subject Sample # 1	0.152	02:25
Breath Volume	3.398	02:25
Air Blank	0.000	02:26
Air Blank	0.000	02:27
Subject Sample # 2	0.142	02:28
Breath Volume	3.546	02:28
Air Blank	0.000	02:29
Control Test	0.079	02:29
Air Blank	0.000	02:30
Diagnostics Check	OK	02:30

Data Download Date/Time: 03/29/2017 15:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUNTER,ROBERT,A
BREATH TEST OPERATOR AGENCY: DeFuniak Springs PD
ARRESTING OFFICER: HUNTER,ROBERT
ARRESTING OFFICER AGENCY: DeFuniak Springs PD
LAST AGENCY INSPECTION: 02/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 00:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:14

SUBJECT'S LAST NAME: LANGFORD

DATE/TIME OF TEST: 03/25/2017 02:17:27

Diagnostics Check	OK	02:07
Air Blank	0.000	02:07
Control Test	0.082	02:08
Air Blank	0.000	02:08
Subject Sample # 1	0.194	02:09
Breath Volume	2.890	02:09
Air Blank	0.000	02:10
Air Blank	0.000	02:12
Subject Sample # 2	0.172	02:12
Breath Volume	1.945	02:12
Air Blank	0.000	02:13
Air Blank	0.000	02:14
Subject Sample # 3	0.181	02:15
Breath Volume	2.867	02:15
Air Blank	0.000	02:16
Control Test	0.081	02:16
Air Blank	0.000	02:17
Diagnostics Check	OK	02:17

Data Download Date/Time: 03/29/2017 15:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PATE,CHARLES,S
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: PATE,C
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 03/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163937
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: ADMIN
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:30

SUBJECT'S LAST NAME: POTTER

DATE/TIME OF TEST: 03/30/2017 17:35:48

Diagnostics Check	OK	17:28
Air Blank	0.000	17:29
Control Test	0.080	17:29
Air Blank	0.000	17:29
Subject Sample # 1	0.145	17:31
Breath Volume	2.214	17:31
Air Blank	0.000	17:32
Air Blank	0.000	17:33
Subject Sample # 2	0.141	17:34
Breath Volume	1.320	17:34
Air Blank	0.000	17:34
Control Test	0.079	17:35
Air Blank	0.000	17:35
Diagnostics Check	OK	17:35

Data Download Date/Time: 04/28/2017 13:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUTTO,KYLE,K
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: EPLEY,CAMERON
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 03/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 19:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:43

SUBJECT'S LAST NAME: MURRAY

DATE/TIME OF TEST: 03/30/2017 20:14:02

Diagnostics Check	OK	20:07	
Air Blank	0.000	20:08	
Control Test	0.080	20:08	
Air Blank	0.000	20:08	
Subject Sample	# 1 0.250	20:09	
Breath Volume	2.257	20:09	
Air Blank	0.000	20:10	
Air Blank	0.000	20:12	
Subject Sample	# 2	20:12	Slope Not Met
Breath Volume	1.082	20:12	
Air Blank	0.000	20:13	
Control Test	0.080	20:13	
Air Blank	0.000	20:13	
Diagnostics Check	OK	20:14	

Data Download Date/Time: 04/28/2017 13:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUTTO,KYLE,K
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: EPLEY,CAMERON
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 03/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 17:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:43

SUBJECT'S LAST NAME: MURRAY

DATE/TIME OF TEST: 03/30/2017 20:24:07

Diagnostics Check	OK	20:17
Air Blank	0.000	20:17
Control Test	0.080	20:18
Air Blank	0.000	20:18
Subject Sample # 1	0.252	20:19
Breath Volume	3.367	20:19
Air Blank	0.000	20:20
Air Blank	0.000	20:22
Subject Sample # 2	0.249	20:22
Breath Volume	1.843	20:22
Air Blank	0.000	20:23
Control Test	0.079	20:23
Air Blank	0.000	20:24
Diagnostics Check	OK	20:24

Data Download Date/Time: 04/28/2017 13:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILEY,DAVID,W
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: WILEY,DAVID
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 02/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: DUI
ARREST TIME: 18:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:58

SUBJECT'S LAST NAME: BULLOCK

DATE/TIME OF TEST: 03/04/2017 19:42:17

Diagnostics Check	OK	19:35
Air Blank	0.000	19:36
Control Test	0.080	19:36
Air Blank	0.000	19:37
Subject Sample # 1	0.252	19:37
Breath Volume	2.406	19:37
Air Blank	0.000	19:38
Air Blank	0.000	19:40
Subject Sample # 2	0.256	19:40
Breath Volume	2.074	19:40
Air Blank	0.000	19:41
Control Test	0.080	19:41
Air Blank	0.000	19:42
Diagnostics Check	OK	19:42

Data Download Date/Time: 03/29/2017 15:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIERNACKI, CHAD, A
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: TICE, HUNTER
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 02/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: DUI
ARREST TIME: 01:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:06

SUBJECT'S LAST NAME: BRIMHALL

DATE/TIME OF TEST: 03/10/2017 02:44:05

Diagnostics Check	OK	02:30	
Air Blank	0.000	02:31	
Control Test	0.079	02:31	
Air Blank	0.000	02:32	
Subject Sample	# 1	02:35	No Sample Provided
Breath Volume	0.000	02:35	
Air Blank	0.000	02:35	
Air Blank	0.000	02:37	
Subject Sample	# 2	0.161	02:39
Breath Volume	1.468	02:39	
Air Blank	0.000	02:39	
Air Blank	0.000	02:41	
Subject Sample	# 3	0.165	02:42
Breath Volume	1.937	02:42	
Air Blank	0.000	02:43	
Control Test	0.076	02:43	
Air Blank	0.000	02:43	
Diagnostics Check	OK	02:44	

Data Download Date/Time: 03/29/2017 15:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHULTZ,JEREMY,J
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: SCHULTZ,JEREMY
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 02/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: DUI
ARREST TIME: 20:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: MORRIS

DATE/TIME OF TEST: 03/12/2017 21:49:12

Diagnostics Check	OK	21:41	
Air Blank	0.000	21:41	
Control Test	0.080	21:41	
Air Blank	0.000	21:42	
Subject Sample	# 1	21:45	No Sample Provided
Breath Volume	0.000	21:45	
Air Blank	0.000	21:45	
Air Blank		21:47	RFI Detect
Air Blank	0.000	21:48	
Control Test	0.079	21:48	
Air Blank	0.000	21:49	
Diagnostics Check	OK	21:49	

Data Download Date/Time: 03/29/2017 15:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHULTZ,JEREMY,J
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: SCHULTZ,JEREMY
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 02/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: DUI
ARREST TIME: 21:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: MORRIS

DATE/TIME OF TEST: 03/12/2017 22:08:03

Diagnostics Check	OK	21:56	
Air Blank	0.000	21:57	
Control Test	0.080	21:57	
Air Blank	0.000	21:57	
Subject Sample	# 1	22:01	No Sample Provided
Breath Volume	0.000	22:01	
Air Blank	0.000	22:01	
Air Blank	0.000	22:03	
Subject Sample	# 2	22:06	No Sample Provided
Breath Volume	0.000	22:06	
Air Blank	0.000	22:07	
Control Test	0.079	22:07	
Air Blank	0.000	22:07	
Diagnostics Check	OK	22:08	

Data Download Date/Time: 03/29/2017 15:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,JONATHAN,B
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: BUDUEN,PABLO
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 02/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: DUI
ARREST TIME: 19:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:27

SUBJECT'S LAST NAME: BRACKETT

DATE/TIME OF TEST: 03/14/2017 20:57:13

Diagnostics Check	OK	20:52	
Air Blank	0.000	20:52	
Control Test	0.079	20:53	
Air Blank	0.000	20:53	
Subject Sample	# 1	20:55	Subject Test Refused
Breath Volume	0.000	20:55	
Air Blank	0.000	20:56	
Control Test	0.081	20:56	
Air Blank	0.000	20:57	
Diagnostics Check	OK	20:57	

Data Download Date/Time: 03/29/2017 15:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIERNACKI,CHAD,A
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: BIERNACKI,CHAD
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 02/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: DUI
ARREST TIME: 23:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:09

SUBJECT'S LAST NAME: ODOM

DATE/TIME OF TEST: 03/18/2017 00:42:44

Diagnostics Check	OK	00:34	
Air Blank	0.000	00:34	
Control Test	0.078	00:35	
Air Blank	0.000	00:35	
Subject Sample	# 1	00:38	No Sample Provided
Breath Volume	0.000	00:38	
Air Blank	0.000	00:39	
Air Blank	0.000	00:41	
Subject Sample	# 2	00:41	Subject Test Refused
Breath Volume	0.000	00:41	
Air Blank	0.000	00:41	
Control Test	0.079	00:42	
Air Blank	0.000	00:42	
Diagnostics Check	OK	00:42	

Data Download Date/Time: 03/29/2017 15:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIERNACKI,CHAD,A
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: BUDUEN,PABLO
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 02/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: DUI
ARREST TIME: 21:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: DE LA CRUZ

DATE/TIME OF TEST: 03/18/2017 22:07:59

Diagnostics Check	OK	22:01
Air Blank	0.000	22:01
Control Test	0.080	22:01
Air Blank	0.000	22:02
Subject Sample # 1	0.221	22:02
Breath Volume	1.621	22:02
Air Blank	0.000	22:03
Air Blank	0.000	22:05
Subject Sample # 2	0.240	22:06
Breath Volume	1.542	22:06
Air Blank	0.000	22:07
Control Test	0.079	22:07
Air Blank	0.000	22:07
Diagnostics Check	OK	22:07

Data Download Date/Time: 03/29/2017 15:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GAYDOS,JONATHAN,A
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: GAYDOS,JONATHAN
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 02/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: DUI
ARREST TIME: 15:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:59

SUBJECT'S LAST NAME: PARKER

DATE/TIME OF TEST: 03/23/2017 16:26:06

Diagnostics Check	OK	16:22	
Air Blank	0.000	16:23	
Control Test	0.078	16:23	
Air Blank	0.000	16:24	
Subject Sample	# 1	16:24	Subject Test Refused
Breath Volume	0.000	16:24	
Air Blank	0.000	16:25	
Control Test	0.079	16:25	
Air Blank	0.000	16:26	
Diagnostics Check	OK	16:26	

Data Download Date/Time: 03/29/2017 15:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIERNACKI,CHAD,A
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: BIERNACKI,CHAD
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 02/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: DUI
ARREST TIME: 02:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:52

SUBJECT'S LAST NAME: JOB

DATE/TIME OF TEST: 03/24/2017 03:20:40

Diagnostics Check	OK	03:12	
Air Blank	0.000	03:12	
Control Test	0.079	03:12	
Air Blank	0.000	03:13	
Subject Sample	# 1	03:16	No Sample Provided
Breath Volume	0.000	03:16	
Air Blank	0.000	03:17	
Air Blank	0.000	03:19	
Subject Sample	# 2	03:19	Subject Test Refused
Breath Volume	0.000	03:19	
Air Blank	0.000	03:19	
Control Test	0.080	03:20	
Air Blank	0.000	03:20	
Diagnostics Check	OK	03:20	

Data Download Date/Time: 03/29/2017 15:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHULTZ,JEREMY,J
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: SCHULTZ,JEREMY
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 02/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: DUI
ARREST TIME: 23:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:29

SUBJECT'S LAST NAME: ;6360101260072162045

DATE/TIME OF TEST: 03/26/2017 00:00:39

Diagnostics Check	OK	23:49	
Air Blank	0.000	23:49	
Control Test	0.079	23:50	
Air Blank	0.000	23:50	
Subject Sample	# 1	23:53	No Sample Provided
Breath Volume	0.000	23:53	
Air Blank	0.000	23:54	
Air Blank	0.000	23:56	
Subject Sample	# 2	23:59	No Sample Provided
Breath Volume	0.000	23:59	
Air Blank	0.000	23:59	
Control Test	0.079	00:00	
Air Blank	0.000	00:00	
Diagnostics Check	OK	00:00	

Data Download Date/Time: 03/29/2017 15:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001313

INSTRUMENT REGISTERED TO: SPRINGFIELD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KING,MIKE,B
BREATH TEST OPERATOR AGENCY: Springfield PD
ARRESTING OFFICER: ROBERTS,BARRY
ARRESTING OFFICER AGENCY: Springfield PD
LAST AGENCY INSPECTION: 02/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 10:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:41

SUBJECT'S LAST NAME: MANN

DATE/TIME OF TEST: 03/29/2017 11:08:25

Diagnostics Check	OK	11:02
Air Blank	0.000	11:02
Control Test	0.078	11:02
Air Blank	0.000	11:03
Subject Sample # 1	0.000	11:03
Breath Volume	4.246	11:03
Air Blank	0.000	11:04
Air Blank	0.000	11:06
Subject Sample # 2	0.000	11:07
Breath Volume	3.671	11:07
Air Blank	0.000	11:07
Control Test	0.079	11:07
Air Blank	0.000	11:08
Diagnostics Check	OK	11:08

Data Download Date/Time: 05/30/2017 14:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YERKES,BECKY,A
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: HENDERSON,JOANN
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 02/03/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: DUI
ARREST TIME: 02:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:36

SUBJECT'S LAST NAME: YERKES

DATE/TIME OF TEST: 03/04/2017 02:58:43

Diagnostics Check	OK	02:56	
Air Blank	0.000	02:57	
Control Test	0.081	02:57	
Air Blank	0.000	02:58	
Subject Sample	# 1	02:58	Sequence Aborted
Breath Volume	0.000	02:58	
Air Blank	0.000	02:58	

Data Download Date/Time: 03/22/2017 14:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,CARL,C
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: HENDERSON,JOANN
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 02/03/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: DUI
ARREST TIME: 02:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:36

SUBJECT'S LAST NAME: YERKES

DATE/TIME OF TEST: 03/04/2017 03:08:38

Diagnostics Check	OK	03:00
Air Blank	0.000	03:01
Control Test	0.081	03:01
Air Blank	0.000	03:02
Subject Sample # 1	0.244	03:03
Breath Volume	2.578	03:03
Air Blank	0.000	03:04
Air Blank	0.000	03:05
Subject Sample # 2	0.233	03:07
Breath Volume	2.367	03:07
Air Blank	0.000	03:07
Control Test	0.080	03:08
Air Blank	0.000	03:08
Diagnostics Check	OK	03:08

Data Download Date/Time: 03/22/2017 14:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POLLOCK,STEVEN,D
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: POLLOCK,STEVEN
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 03/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: DUI
ARREST TIME: 02:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:13

SUBJECT'S LAST NAME: HICKMAN

DATE/TIME OF TEST: 03/23/2017 03:41:04

Diagnostics Check	OK	03:34
Air Blank	0.000	03:35
Control Test	0.081	03:35
Air Blank	0.000	03:36
Subject Sample # 1	0.254	03:36
Breath Volume	3.207	03:36
Air Blank	0.000	03:37
Air Blank	0.000	03:39
Subject Sample # 2	0.245	03:39
Breath Volume	3.417	03:39
Air Blank	0.000	03:40
Control Test	0.080	03:40
Air Blank	0.000	03:40
Diagnostics Check	OK	03:41

Data Download Date/Time: 04/17/2017 13:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUEDE,KAYLA,M
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: RUEDE,KAYLA
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 02/28/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: BENEDETTO

DATE/TIME OF TEST: 03/06/2017 04:44:33

Diagnostics Check	OK	04:38
Air Blank	0.000	04:38
Control Test	0.081	04:38
Air Blank	0.000	04:39
Subject Sample # 1	0.020	04:39
Breath Volume	1.140	04:39
Air Blank	0.000	04:40
Air Blank	0.000	04:42
Subject Sample # 2	0.018	04:43
Breath Volume	1.390	04:43
Air Blank	0.000	04:43
Control Test	0.081	04:43
Air Blank	0.000	04:44
Diagnostics Check	OK	04:44

Data Download Date/Time: 03/28/2017 13:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEWIS,JEREMY,A
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: ZAMOYSKI,ADRIAN
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 02/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: GALLANDER

DATE/TIME OF TEST: 03/09/2017 22:17:39

Diagnostics Check	OK	22:11
Air Blank	0.000	22:11
Control Test	0.081	22:11
Air Blank	0.000	22:12
Subject Sample # 1	0.078	22:13
Breath Volume	1.800	22:13
Air Blank	0.000	22:13
Air Blank	0.000	22:15
Subject Sample # 2	0.071	22:16
Breath Volume	1.574	22:16
Air Blank	0.000	22:16
Control Test	0.081	22:17
Air Blank	0.000	22:17
Diagnostics Check	OK	22:17

Data Download Date/Time: 03/28/2017 13:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEWIS,JEREMY,A
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: PAPALEO,DANIEL
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 02/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: RAMSEY

DATE/TIME OF TEST: 03/11/2017 01:51:52

Diagnostics Check	OK	01:45
Air Blank	0.000	01:46
Control Test	0.081	01:46
Air Blank	0.000	01:47
Subject Sample # 1	0.166	01:47
Breath Volume	1.726	01:47
Air Blank	0.000	01:48
Air Blank	0.000	01:49
Subject Sample # 2	0.161	01:50
Breath Volume	1.332	01:50
Air Blank	0.000	01:50
Control Test	0.080	01:51
Air Blank	0.000	01:51
Diagnostics Check	OK	01:51

Data Download Date/Time: 03/28/2017 13:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUEDE,KAYLA,M
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: RUEDE,KAYLA
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 02/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:06

SUBJECT'S LAST NAME: GARLAND

DATE/TIME OF TEST: 03/12/2017 23:43:31

Diagnostics Check	OK	23:34
Air Blank	0.000	23:34
Control Test	0.080	23:35
Air Blank	0.000	23:35
Subject Sample # 1	0.172	23:35
Breath Volume	1.652	23:35
Air Blank	0.000	23:36
Air Blank	0.000	23:38
Subject Sample # 2	0.195	23:39
Breath Volume	1.800	23:39
Air Blank	0.000	23:39
Air Blank	0.000	23:41
Subject Sample # 3	0.171	23:41
Breath Volume	1.628	23:41
Air Blank	0.000	23:42
Control Test	0.080	23:42
Air Blank	0.000	23:43
Diagnostics Check	OK	23:43

Data Download Date/Time: 03/28/2017 13:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PETO,JOSHUA,W
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: RODRIGUEZ,RAFAEL
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 02/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 18:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:05

SUBJECT'S LAST NAME: LEONE

DATE/TIME OF TEST: 03/15/2017 19:39:25

Diagnostics Check	OK	19:32
Air Blank	0.000	19:33
Control Test	0.082	19:33
Air Blank	0.000	19:34
Subject Sample # 1	0.037	19:34
Breath Volume	2.828	19:34
Air Blank	0.000	19:35
Air Blank	0.000	19:37
Subject Sample # 2	0.035	19:37
Breath Volume	1.687	19:37
Air Blank	0.000	19:38
Control Test	0.082	19:38
Air Blank	0.000	19:39
Diagnostics Check	OK	19:39

Data Download Date/Time: 03/28/2017 13:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEINKRAUS,ROBERT,W
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: ROMANI,ANTHONY
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 02/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 13:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:00

SUBJECT'S LAST NAME: ABEL

DATE/TIME OF TEST: 03/26/2017 14:33:57

Diagnostics Check	OK	14:26
Air Blank	0.000	14:26
Control Test	0.081	14:27
Air Blank	0.000	14:27
Subject Sample # 1	0.069	14:29
Breath Volume	1.496	14:29
Air Blank	0.000	14:29
Air Blank	0.000	14:31
Subject Sample # 2	0.079	14:32
Breath Volume	2.675	14:32
Air Blank	0.000	14:33
Control Test	0.080	14:33
Air Blank	0.000	14:33
Diagnostics Check	OK	14:33

Data Download Date/Time: 03/28/2017 13:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUZMAN,MELISSA,M
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: NOONAN,JAMES
ARRESTING OFFICER AGENCY: Sebastian PD
LAST AGENCY INSPECTION: 02/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 23:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: GRAHAM

DATE/TIME OF TEST: 03/01/2017 01:13:22

Diagnostics Check	OK	01:10	
Air Blank	0.000	01:10	
Control Test	0.081	01:11	
Air Blank	0.000	01:11	
Subject Sample	# 1	01:12	Subject Test Refused
Breath Volume	0.000	01:12	
Air Blank	0.000	01:12	
Control Test	0.081	01:12	
Air Blank	0.000	01:13	
Diagnostics Check	OK	01:13	

Data Download Date/Time: 03/05/2017 06:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHIMENTI,RICHARD,J
BREATH TEST OPERATOR AGENCY: Vero Beach PD
ARRESTING OFFICER: CLAWSON,NOAH
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 02/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 02:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: GHETTI-ANDERSEN

DATE/TIME OF TEST: 03/01/2017 03:14:45

Diagnostics Check	OK	03:11	
Air Blank	0.000	03:12	
Control Test	0.080	03:12	
Air Blank	0.000	03:13	
Subject Sample	# 1	03:13	Subject Test Refused
Breath Volume	0.000	03:13	
Air Blank	0.000	03:13	
Control Test	0.081	03:14	
Air Blank	0.000	03:14	
Diagnostics Check	OK	03:14	

Data Download Date/Time: 03/05/2017 06:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SWINDELL,CHRISTOPHER,E
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: FLORES,DAVID
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 02/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:26

SUBJECT'S LAST NAME: EXCLUSA

DATE/TIME OF TEST: 03/02/2017 01:51:31

Diagnostics Check	OK	01:48	
Air Blank	0.000	01:49	
Control Test	0.081	01:49	
Air Blank	0.000	01:49	
Subject Sample	# 1	01:50	Subject Test Refused
Breath Volume	0.000	01:50	
Air Blank	0.000	01:50	
Control Test	0.081	01:50	
Air Blank	0.000	01:51	
Diagnostics Check	OK	01:51	

Data Download Date/Time: 03/05/2017 06:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSADO,L,M
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: TAYLOR,K
ARRESTING OFFICER AGENCY: Fellsmere PD
LAST AGENCY INSPECTION: 02/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 22:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:44

SUBJECT'S LAST NAME: VILLA

DATE/TIME OF TEST: 03/04/2017 01:13:47

Diagnostics Check	OK	01:07
Air Blank	0.000	01:07
Control Test	0.082	01:08
Air Blank	0.000	01:08
Subject Sample # 1	0.311	01:09
Breath Volume	3.195	01:09
Air Blank	0.000	01:09
Air Blank	0.000	01:11
Subject Sample # 2	0.308	01:12
Breath Volume	3.312	01:12
Air Blank	0.000	01:12
Control Test	0.081	01:13
Air Blank	0.000	01:13
Diagnostics Check	OK	01:13

Data Download Date/Time: 03/05/2017 06:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STOVALL,C,J
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: JAWORSKI,K
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 02/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 16:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:05

SUBJECT'S LAST NAME: JACOBSON

DATE/TIME OF TEST: 03/04/2017 17:41:51

Diagnostics Check	OK	17:39	
Air Blank	0.000	17:39	
Control Test	0.080	17:39	
Air Blank	0.000	17:40	
Subject Sample	# 1	17:40	Subject Test Refused
Breath Volume	0.000	17:40	
Air Blank	0.000	17:40	
Control Test	0.081	17:41	
Air Blank	0.000	17:41	
Diagnostics Check	OK	17:41	

Data Download Date/Time: 03/05/2017 06:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOMINGUEZ, OSCAR,
BREATH TEST OPERATOR AGENCY: Vero Beach PD
ARRESTING OFFICER: WATERS, THOMAS
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 02/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 02:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:41

SUBJECT'S LAST NAME: ESQUIBEL

DATE/TIME OF TEST: 03/05/2017 03:07:56

Diagnostics Check	OK	03:01
Air Blank	0.000	03:02
Control Test	0.081	03:02
Air Blank	0.000	03:02
Subject Sample # 1	0.234	03:03
Breath Volume	2.074	03:03
Air Blank	0.000	03:03
Air Blank	0.000	03:05
Subject Sample # 2	0.229	03:06
Breath Volume	1.261	03:06
Air Blank	0.000	03:07
Control Test	0.080	03:07
Air Blank	0.000	03:07
Diagnostics Check	OK	03:07

Data Download Date/Time: 03/05/2017 06:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROOKS, MATTHEW, W
BREATH TEST OPERATOR AGENCY: Vero Beach PD
ARRESTING OFFICER: BROOKS, MATTHEW
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 02/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 03:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:12

SUBJECT'S LAST NAME: RAMIREZ

DATE/TIME OF TEST: 03/05/2017 04:52:07

Diagnostics Check	OK	04:39
Air Blank	0.000	04:39
Control Test	0.081	04:39
Air Blank	0.000	04:40
Subject Sample # 1	0.256	04:41
Breath Volume	2.039	04:41
Air Blank	0.000	04:42
Air Blank	0.000	04:44
Subject Sample # 2	0.224	04:47
Breath Volume	1.542	04:47
Air Blank	0.000	04:48
Air Blank	0.000	04:49
Subject Sample # 3	0.238	04:50
Breath Volume	1.585	04:50
Air Blank	0.000	04:51
Control Test	0.079	04:51
Air Blank	0.000	04:52
Diagnostics Check	OK	04:52

Data Download Date/Time: 03/05/2017 06:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,LEE,A
BREATH TEST OPERATOR AGENCY: Vero Beach PD
ARRESTING OFFICER: EVANS,LEE
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 03/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 05:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:48

SUBJECT'S LAST NAME: HAMILTON

DATE/TIME OF TEST: 03/05/2017 07:11:45

Diagnostics Check	OK	07:08	
Air Blank	0.000	07:09	
Control Test	0.082	07:09	
Air Blank	0.000	07:10	
Subject Sample	# 1	07:10	Subject Test Refused
Breath Volume	0.000	07:10	
Air Blank	0.000	07:10	
Control Test	0.081	07:11	
Air Blank	0.000	07:11	
Diagnostics Check	OK	07:11	

Data Download Date/Time: 04/16/2017 05:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STOVALL,C,J
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: BELLEFLEUR,C
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 03/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 21:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: PETTICREW

DATE/TIME OF TEST: 03/05/2017 22:56:12

Diagnostics Check	OK	22:53	
Air Blank	0.000	22:53	
Control Test	0.082	22:54	
Air Blank	0.000	22:54	
Subject Sample	# 1	22:54	Subject Test Refused
Breath Volume	0.000	22:54	
Air Blank	0.000	22:55	
Control Test	0.081	22:55	
Air Blank	0.000	22:56	
Diagnostics Check	OK	22:56	

Data Download Date/Time: 04/16/2017 05:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EGGERS,RYAN,K
BREATH TEST OPERATOR AGENCY: Vero Beach PD
ARRESTING OFFICER: EGGERS,RYAN
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 03/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: LESLIE

DATE/TIME OF TEST: 03/06/2017 02:26:16

Diagnostics Check	OK	02:23	
Air Blank	0.000	02:23	
Control Test	0.081	02:24	
Air Blank	0.000	02:24	
Subject Sample	# 1	02:24	Subject Test Refused
Breath Volume	0.000	02:24	
Air Blank	0.000	02:25	
Control Test	0.081	02:25	
Air Blank	0.000	02:26	
Diagnostics Check	OK	02:26	

Data Download Date/Time: 04/16/2017 05:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EGGERS,RYAN,K
BREATH TEST OPERATOR AGENCY: Vero Beach PD
ARRESTING OFFICER: EGGERS,RYAN
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 03/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 04:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:55

SUBJECT'S LAST NAME: CAPAS

DATE/TIME OF TEST: 03/06/2017 05:21:46

Diagnostics Check	OK	05:18	
Air Blank	0.000	05:19	
Control Test	0.082	05:19	
Air Blank	0.000	05:20	
Subject Sample	# 1	05:20	Subject Test Refused
Breath Volume	0.000	05:20	
Air Blank	0.000	05:20	
Control Test	0.081	05:21	
Air Blank	0.000	05:21	
Diagnostics Check	OK	05:21	

Data Download Date/Time: 04/16/2017 05:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PLUMB,GREGORY,T
BREATH TEST OPERATOR AGENCY: Vero Beach PD
ARRESTING OFFICER: PLUMB,GREGORY
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 03/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 01:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: HOWELL

DATE/TIME OF TEST: 03/07/2017 01:43:21

Diagnostics Check	OK	01:40
Air Blank	0.000	01:40
Control Test	0.081	01:41
Air Blank	0.000	01:41
Subject Sample # 1	0.153	01:42
Breath Volume	3.601	01:42
Air Blank	0.000	01:42
Air Blank	0.000	01:43

Data Download Date/Time: 04/16/2017 05:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PLUMB,GREGORY,T
BREATH TEST OPERATOR AGENCY: Vero Beach PD
ARRESTING OFFICER: PLUMB,GREGORY
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 03/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 01:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: HOWELL

DATE/TIME OF TEST: 03/07/2017 01:50:59

Diagnostics Check	OK	01:44
Air Blank	0.000	01:45
Control Test	0.080	01:45
Air Blank	0.000	01:46
Subject Sample # 1	0.148	01:46
Breath Volume	2.535	01:46
Air Blank	0.000	01:47
Air Blank	0.000	01:49
Subject Sample # 2	0.149	01:49
Breath Volume	3.078	01:49
Air Blank	0.000	01:50
Control Test	0.080	01:50
Air Blank	0.000	01:50
Diagnostics Check	OK	01:50

Data Download Date/Time: 04/16/2017 05:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SWINDELL,CHRISTOPHER,E
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: FLORES,DAVID
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 03/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 23:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: SCARFO

DATE/TIME OF TEST: 03/08/2017 00:37:16

Diagnostics Check	OK	00:34	
Air Blank	0.000	00:34	
Control Test	0.081	00:35	
Air Blank	0.000	00:35	
Subject Sample	# 1	00:35	Subject Test Refused
Breath Volume	0.000	00:35	
Air Blank	0.000	00:36	
Control Test	0.081	00:36	
Air Blank	0.000	00:37	
Diagnostics Check	OK	00:37	

Data Download Date/Time: 04/16/2017 05:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SWINDELL,CHRISTOPHER,E
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: FLORES,DAVID
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 03/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 03:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:23

SUBJECT'S LAST NAME: NALEWAISKI

DATE/TIME OF TEST: 03/08/2017 03:51:22

Diagnostics Check	OK	03:48	
Air Blank	0.000	03:48	
Control Test	0.082	03:49	
Air Blank	0.000	03:49	
Subject Sample	# 1	03:50	Subject Test Refused
Breath Volume	0.000	03:50	
Air Blank	0.000	03:50	
Control Test	0.082	03:50	
Air Blank	0.000	03:51	
Diagnostics Check	OK	03:51	

Data Download Date/Time: 04/16/2017 05:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SWINDELL,CHRISTOPHER,E
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: THIMMER,CRAIG
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 03/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 23:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: LUCE

DATE/TIME OF TEST: 03/09/2017 00:26:49

Diagnostics Check	OK	00:19
Air Blank	0.000	00:20
Control Test	0.081	00:20
Air Blank	0.000	00:20
Subject Sample # 1	0.170	00:22
Breath Volume	2.050	00:22
Air Blank	0.000	00:22
Air Blank	0.000	00:24
Subject Sample # 2	0.170	00:25
Breath Volume	1.949	00:25
Air Blank	0.000	00:25
Control Test	0.082	00:26
Air Blank	0.000	00:26
Diagnostics Check	OK	00:26

Data Download Date/Time: 04/16/2017 05:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUZMAN,MELISSA,M
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: SHARPETA,CHAD
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 03/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 00:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:01

SUBJECT'S LAST NAME: RICAFORT

DATE/TIME OF TEST: 03/10/2017 03:27:35

Diagnostics Check	OK	03:21
Air Blank	0.000	03:22
Control Test	0.081	03:22
Air Blank	0.000	03:22
Subject Sample # 1	0.150	03:23
Breath Volume	1.960	03:23
Air Blank	0.000	03:23
Air Blank	0.000	03:25
Subject Sample # 2	0.148	03:26
Breath Volume	2.343	03:26
Air Blank	0.000	03:26
Control Test	0.080	03:27
Air Blank	0.000	03:27
Diagnostics Check	OK	03:27

Data Download Date/Time: 04/16/2017 05:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOLSKI,JEFFREY,R
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: JAWORSKI,KEVIN
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 03/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 16:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:51

SUBJECT'S LAST NAME: POLLICK

DATE/TIME OF TEST: 03/10/2017 17:24:13

Diagnostics Check	OK	17:21	
Air Blank	0.000	17:21	
Control Test	0.082	17:22	
Air Blank	0.000	17:22	
Subject Sample	# 1	17:22	Subject Test Refused
Breath Volume	0.000	17:22	
Air Blank	0.000	17:23	
Control Test	0.081	17:23	
Air Blank	0.000	17:24	
Diagnostics Check	OK	17:24	

Data Download Date/Time: 04/16/2017 05:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOERGER,SHANE,A
BREATH TEST OPERATOR AGENCY: Vero Beach PD
ARRESTING OFFICER: JOERGER,SHANE
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 03/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 02:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: CASTELLANOS

DATE/TIME OF TEST: 03/11/2017 03:35:43

Diagnostics Check	OK	03:29
Air Blank	0.000	03:29
Control Test	0.081	03:30
Air Blank	0.000	03:30
Subject Sample # 1	0.020	03:31
Breath Volume	1.937	03:31
Air Blank	0.000	03:31
Air Blank	0.000	03:33
Subject Sample # 2	0.018	03:34
Breath Volume	1.167	03:34
Air Blank	0.000	03:34
Control Test	0.080	03:35
Air Blank	0.000	03:35
Diagnostics Check	OK	03:35

Data Download Date/Time: 04/16/2017 05:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOLSKI,JEFFREY,R
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: PEREZ,C
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 03/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 18:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:04

SUBJECT'S LAST NAME: CARMAN

DATE/TIME OF TEST: 03/11/2017 21:26:52

Diagnostics Check	OK	21:24	
Air Blank	0.000	21:24	
Control Test	0.081	21:25	
Air Blank	0.000	21:25	
Subject Sample	# 1	21:26	RFI Detect
Breath Volume	0.574	21:26	
Air Blank		21:26	RFI Detect
Air Blank	0.000	21:26	

Data Download Date/Time: 04/16/2017 05:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOLSKI,JEFFREY,R
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: PEREZ,C
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 03/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 18:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:04

SUBJECT'S LAST NAME: CARMAN

DATE/TIME OF TEST: 03/11/2017 21:35:26

Diagnostics Check	OK	21:29
Air Blank	0.000	21:29
Control Test	0.081	21:29
Air Blank	0.000	21:30
Subject Sample # 1	0.180	21:30
Breath Volume	1.687	21:30
Air Blank	0.000	21:31
Air Blank	0.000	21:33
Subject Sample # 2	0.173	21:33
Breath Volume	1.343	21:33
Air Blank	0.000	21:34
Control Test	0.081	21:34
Air Blank	0.000	21:35
Diagnostics Check	OK	21:35

Data Download Date/Time: 04/16/2017 05:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUZMAN,MELISSA,M
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: RUIZ,MICHAEL
ARRESTING OFFICER AGENCY: Fellsmere PD
LAST AGENCY INSPECTION: 03/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 04:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:44

SUBJECT'S LAST NAME: REYES

DATE/TIME OF TEST: 03/12/2017 05:07:22

Diagnostics Check	OK	05:04	
Air Blank	0.000	05:05	
Control Test	0.081	05:05	
Air Blank	0.000	05:05	
Subject Sample	# 1	05:06	Subject Test Refused
Breath Volume	0.000	05:06	
Air Blank	0.000	05:06	
Control Test	0.081	05:06	
Air Blank	0.000	05:07	
Diagnostics Check	OK	05:07	

Data Download Date/Time: 04/16/2017 05:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PLUMB,GREGORY,T
BREATH TEST OPERATOR AGENCY: Vero Beach PD
ARRESTING OFFICER: PLUMB,GREGORY
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 03/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:51

SUBJECT'S LAST NAME: DIPAOLOA

DATE/TIME OF TEST: 03/13/2017 02:19:11

Diagnostics Check	OK	02:11
Air Blank	0.000	02:12
Control Test	0.081	02:12
Air Blank	0.000	02:12
Subject Sample # 1	0.087	02:14
Breath Volume	1.906	02:14
Air Blank	0.000	02:15
Air Blank	0.000	02:16
Subject Sample # 2	0.087	02:17
Breath Volume	2.203	02:17
Air Blank	0.000	02:18
Control Test	0.080	02:18
Air Blank	0.000	02:19
Diagnostics Check	OK	02:19

Data Download Date/Time: 04/16/2017 05:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOERGER,SHANE,A
BREATH TEST OPERATOR AGENCY: Vero Beach PD
ARRESTING OFFICER: WILLIS,JEREMIAH
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 03/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: BLOOMER

DATE/TIME OF TEST: 03/17/2017 01:02:35

Diagnostics Check	OK	00:59	
Air Blank	0.000	00:59	
Control Test	0.081	01:00	
Air Blank	0.000	01:00	
Subject Sample	# 1	01:01	Subject Test Refused
Breath Volume	0.000	01:01	
Air Blank	0.000	01:01	
Control Test	0.081	01:01	
Air Blank	0.000	01:02	
Diagnostics Check	OK	01:02	

Data Download Date/Time: 04/16/2017 05:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSADO,L,M
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: RUIZ,M
ARRESTING OFFICER AGENCY: Fellsmere PD
LAST AGENCY INSPECTION: 03/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 01:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:39

SUBJECT'S LAST NAME: BURT

DATE/TIME OF TEST: 03/17/2017 03:06:19

Diagnostics Check	OK	02:59
Air Blank	0.000	03:00
Control Test	0.081	03:00
Air Blank	0.000	03:01
Subject Sample # 1	0.190	03:01
Breath Volume	1.929	03:01
Air Blank	0.000	03:02
Air Blank	0.000	03:04
Subject Sample # 2	0.170	03:04
Breath Volume	1.160	03:04
Air Blank	0.000	03:05
Control Test	0.081	03:05
Air Blank	0.000	03:06
Diagnostics Check	OK	03:06

Data Download Date/Time: 04/16/2017 05:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOLSKI,JEFFREY,R
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: FARINACCI,DYLAN
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 03/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 23:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: KOZLOWSKI

DATE/TIME OF TEST: 03/18/2017 23:46:06

Diagnostics Check	OK	23:40	
Air Blank	0.000	23:40	
Control Test	0.082	23:41	
Air Blank	0.000	23:41	
Subject Sample # 1	0.123	23:42	
Breath Volume	1.597	23:42	
Air Blank	0.000	23:42	
Air Blank	0.000	23:44	
Subject Sample # 2	0.119	23:45	
Breath Volume	1.378	23:45	
Air Blank		23:45	RFI Detect
Air Blank	0.000	23:46	

Data Download Date/Time: 04/16/2017 05:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EGGERS,RYAN,K
BREATH TEST OPERATOR AGENCY: Vero Beach PD
ARRESTING OFFICER: EGGERS,RYAN
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 03/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 02:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: RODRIQUEZ

DATE/TIME OF TEST: 03/19/2017 02:49:20

Diagnostics Check	OK	02:41
Air Blank	0.000	02:42
Control Test	0.081	02:42
Air Blank	0.000	02:43
Subject Sample # 1	0.160	02:44
Breath Volume	1.179	02:44
Air Blank	0.000	02:44
Air Blank	0.000	02:46
Subject Sample # 2	0.156	02:47
Breath Volume	1.433	02:47
Air Blank	0.000	02:48
Control Test	0.081	02:48
Air Blank	0.000	02:49
Diagnostics Check	OK	02:49

Data Download Date/Time: 04/16/2017 05:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHIMENTI,RICHARD,J
BREATH TEST OPERATOR AGENCY: Vero Beach PD
ARRESTING OFFICER: SMITH,LAURA
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 03/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 03:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: BATTLE

DATE/TIME OF TEST: 03/19/2017 03:46:38

Diagnostics Check	OK	03:39
Air Blank	0.000	03:40
Control Test	0.081	03:40
Air Blank	0.000	03:41
Subject Sample # 1	0.153	03:42
Breath Volume	1.457	03:42
Air Blank	0.000	03:42
Air Blank	0.000	03:44
Subject Sample # 2	0.154	03:45
Breath Volume	1.703	03:45
Air Blank	0.000	03:45
Control Test	0.081	03:46
Air Blank	0.000	03:46
Diagnostics Check	OK	03:46

Data Download Date/Time: 04/16/2017 05:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STOVALL,C,J
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: LOZADA,J
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 03/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 18:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: KLUCINEC

DATE/TIME OF TEST: 03/19/2017 21:50:27

Diagnostics Check	OK	21:47	
Air Blank	0.000	21:48	
Control Test	0.081	21:48	
Air Blank	0.000	21:48	
Subject Sample	# 1	21:49	Subject Test Refused
Breath Volume	0.000	21:49	
Air Blank	0.000	21:49	
Control Test	0.081	21:49	
Air Blank	0.000	21:50	
Diagnostics Check	OK	21:50	

Data Download Date/Time: 04/16/2017 05:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EGGERS,RYAN,K
BREATH TEST OPERATOR AGENCY: Vero Beach PD
ARRESTING OFFICER: EGGERS,RYAN
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 03/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 22:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: CAPAS

DATE/TIME OF TEST: 03/19/2017 22:50:54

Diagnostics Check	OK	22:43	
Air Blank	0.000	22:44	
Control Test	0.082	22:44	
Air Blank	0.000	22:45	
Subject Sample # 1	0.109	22:46	
Breath Volume	1.757	22:46	
Air Blank	0.000	22:46	
Air Blank	0.000	22:48	
Subject Sample # 2		22:49	Interferent Detect
Breath Volume	2.433	22:49	
Air Blank	0.000	22:50	
Control Test	0.081	22:50	Interferent Detect
Air Blank	0.000	22:50	

Data Download Date/Time: 04/16/2017 05:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EGGERS,RYAN,K
BREATH TEST OPERATOR AGENCY: Vero Beach PD
ARRESTING OFFICER: EGGERS,RYAN
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 03/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 22:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: CAPAS

DATE/TIME OF TEST: 03/19/2017 22:58:03

Diagnostics Check	OK	22:51	
Air Blank	0.000	22:52	
Control Test	0.081	22:52	
Air Blank	0.000	22:53	
Subject Sample	# 1 0.114	22:53	
Breath Volume	3.273	22:53	
Air Blank	0.000	22:54	
Air Blank	0.000	22:56	
Subject Sample	# 2	22:56	Subject Test Refused
Breath Volume	0.000	22:56	
Air Blank	0.000	22:57	
Control Test	0.082	22:57	
Air Blank	0.000	22:57	
Diagnostics Check	OK	22:58	

Data Download Date/Time: 04/16/2017 05:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EGGERS,RYAN,K
BREATH TEST OPERATOR AGENCY: Vero Beach PD
ARRESTING OFFICER: EDER,KYLE
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 03/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: KRAIPRASIT

DATE/TIME OF TEST: 03/21/2017 02:08:43

Diagnostics Check	OK	01:56	
Air Blank	0.000	01:56	
Control Test	0.081	01:57	
Air Blank	0.000	01:57	
Subject Sample	# 1 0.169	01:58	
Breath Volume	1.125	01:58	
Air Blank	0.000	01:59	
Air Blank	0.000	02:01	
Subject Sample	# 2 0.156	02:04	Volume Not Met
Breath Volume	1.070	02:04	
Air Blank	0.000	02:04	
Air Blank	0.000	02:06	
Subject Sample	# 3 0.170	02:07	
Breath Volume	1.410	02:07	
Air Blank	0.000	02:07	
Control Test	0.081	02:08	
Air Blank	0.000	02:08	
Diagnostics Check	OK	02:08	

Data Download Date/Time: 04/16/2017 05:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,QUINTON,T
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: GILLETTE,JASON
ARRESTING OFFICER AGENCY: Sebastian PD
LAST AGENCY INSPECTION: 03/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 22:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: GOFF

DATE/TIME OF TEST: 03/21/2017 23:07:38

Diagnostics Check	OK	23:04	
Air Blank	0.000	23:05	
Control Test	0.081	23:05	
Air Blank	0.000	23:05	
Subject Sample	# 1	23:06	RFI Detect
Breath Volume	0.000	23:06	
Air Blank	0.000	23:06	
Control Test	0.081	23:07	
Air Blank	0.000	23:07	
Diagnostics Check	OK	23:07	

Data Download Date/Time: 04/16/2017 05:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,Q,T
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: GILLETTE,JASON
ARRESTING OFFICER AGENCY: Sebastian PD
LAST AGENCY INSPECTION: 03/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 22:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: GOFF

DATE/TIME OF TEST: 03/21/2017 23:10:47

Diagnostics Check	OK	23:10	
Air Blank		23:10	RFI Detect
Air Blank	0.000	23:10	

Data Download Date/Time: 04/16/2017 05:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUZMAN,MELISSA,M
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: FLORES,DAVID
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 03/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:27

SUBJECT'S LAST NAME: BEACH

DATE/TIME OF TEST: 03/22/2017 02:54:54

Diagnostics Check	OK	02:47
Air Blank	0.000	02:48
Control Test	0.081	02:48
Air Blank	0.000	02:49
Subject Sample # 1	0.133	02:50
Breath Volume	2.050	02:50
Air Blank	0.000	02:50
Air Blank	0.000	02:52
Subject Sample # 2	0.140	02:53
Breath Volume	1.476	02:53
Air Blank	0.000	02:54
Control Test	0.081	02:54
Air Blank	0.000	02:54
Diagnostics Check	OK	02:54

Data Download Date/Time: 04/16/2017 05:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROOKS, MATTHEW, W
BREATH TEST OPERATOR AGENCY: Vero Beach PD
ARRESTING OFFICER: BROOKS, MATTHEW
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 03/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 18:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:17

SUBJECT'S LAST NAME: HOLMES

DATE/TIME OF TEST: 03/23/2017 21:46:03

Diagnostics Check	OK	21:38
Air Blank	0.000	21:38
Control Test	0.081	21:39
Air Blank	0.000	21:39
Subject Sample # 1	0.211	21:41
Breath Volume	2.023	21:41
Air Blank	0.000	21:42
Air Blank	0.000	21:44
Subject Sample # 2	0.208	21:44
Breath Volume	1.359	21:44
Air Blank	0.000	21:45
Control Test	0.080	21:45
Air Blank	0.000	21:45
Diagnostics Check	OK	21:46

Data Download Date/Time: 04/16/2017 05:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STOVALL,C,J
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: SHARPETA,C
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 03/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 21:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: WOLFROM

DATE/TIME OF TEST: 03/23/2017 22:09:41

Diagnostics Check	OK	22:03
Air Blank	0.000	22:03
Control Test	0.081	22:04
Air Blank	0.000	22:04
Subject Sample # 1	0.318	22:05
Breath Volume	3.324	22:05
Air Blank	0.000	22:05
Air Blank	0.000	22:07
Subject Sample # 2	0.308	22:08
Breath Volume	2.488	22:08
Air Blank	0.000	22:08
Control Test	0.080	22:09
Air Blank	0.000	22:09
Diagnostics Check	OK	22:09

Data Download Date/Time: 04/16/2017 05:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUZMAN,MELISSA,M
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: SINGH,TEGPREET
ARRESTING OFFICER AGENCY: Sebastian PD
LAST AGENCY INSPECTION: 03/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 00:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: CADENA BOCANEGRA

DATE/TIME OF TEST: 03/25/2017 01:12:17

Diagnostics Check	OK	01:06
Air Blank	0.000	01:06
Control Test	0.082	01:06
Air Blank	0.000	01:07
Subject Sample # 1	0.155	01:07
Breath Volume	2.894	01:07
Air Blank	0.000	01:08
Air Blank	0.000	01:10
Subject Sample # 2	0.157	01:10
Breath Volume	2.894	01:10
Air Blank	0.000	01:11
Control Test	0.081	01:11
Air Blank	0.000	01:12
Diagnostics Check	OK	01:12

Data Download Date/Time: 04/16/2017 05:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMILTON,DEVIN,C
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: RUIZ,MICHAEL
ARRESTING OFFICER AGENCY: Fellsmere PD
LAST AGENCY INSPECTION: 03/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 22:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: DELGADO

DATE/TIME OF TEST: 03/26/2017 00:46:20

Diagnostics Check	OK	00:38
Air Blank	0.000	00:38
Control Test	0.080	00:39
Air Blank	0.000	00:39
Subject Sample # 1	0.205	00:41
Breath Volume	1.542	00:41
Air Blank	0.000	00:42
Air Blank	0.000	00:44
Subject Sample # 2	0.204	00:44
Breath Volume	2.464	00:44
Air Blank	0.000	00:45
Control Test	0.081	00:45
Air Blank	0.000	00:46
Diagnostics Check	OK	00:46

Data Download Date/Time: 04/16/2017 05:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STOVALL,C,J
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: RUIZ,M
ARRESTING OFFICER AGENCY: Fellsmere PD
LAST AGENCY INSPECTION: 03/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 22:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:17

SUBJECT'S LAST NAME: SANCHEZ

DATE/TIME OF TEST: 03/26/2017 23:43:07

Diagnostics Check	OK	23:37	
Air Blank	0.000	23:37	
Control Test	0.081	23:38	
Air Blank	0.000	23:38	
Subject Sample	# 1 0.085	23:39	
Breath Volume	2.046	23:39	
Air Blank	0.000	23:39	
Air Blank	0.000	23:41	
Subject Sample	# 2	23:41	RFI Detect
Breath Volume	0.000	23:41	
Air Blank	0.000	23:42	
Control Test	0.081	23:42	
Air Blank	0.000	23:43	
Diagnostics Check	OK	23:43	

Data Download Date/Time: 04/16/2017 05:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STOVALL,C,J
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: RUIZ,M
ARRESTING OFFICER AGENCY: Fellsmere PD
LAST AGENCY INSPECTION: 03/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 22:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:17

SUBJECT'S LAST NAME: SANCHEZ

DATE/TIME OF TEST: 03/26/2017 23:51:41

Diagnostics Check	OK	23:45
Air Blank	0.000	23:45
Control Test	0.081	23:45
Air Blank	0.000	23:46
Subject Sample # 1	0.084	23:46
Breath Volume	2.253	23:46
Air Blank	0.000	23:47
Air Blank	0.000	23:49
Subject Sample # 2	0.086	23:50
Breath Volume	2.824	23:50
Air Blank	0.000	23:50
Control Test	0.081	23:51
Air Blank	0.000	23:51
Diagnostics Check	OK	23:51

Data Download Date/Time: 04/16/2017 05:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,QUINTON,T
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: GILLETTE,JASON
ARRESTING OFFICER AGENCY: Sebastian PD
LAST AGENCY INSPECTION: 03/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 684131
DRY GAS EXPIRY DATE: 01/23/2020
VIOLATION CODE: DUI
ARREST TIME: 22:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: GOFF

DATE/TIME OF TEST: 03/21/2017 23:26:39

Diagnostics Check	OK	23:20
Air Blank	0.000	23:20
Control Test	0.080	23:21
Air Blank	0.000	23:21
Subject Sample # 1	0.255	23:22
Breath Volume	2.660	23:22
Air Blank	0.000	23:22
Air Blank	0.000	23:24
Subject Sample # 2	0.260	23:25
Breath Volume	2.773	23:25
Air Blank	0.000	23:25
Control Test	0.079	23:26
Air Blank	0.000	23:26
Diagnostics Check	OK	23:26

Data Download Date/Time: 04/16/2017 05:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001341

INSTRUMENT REGISTERED TO: DESOTO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HORTON, EDMOND, W
BREATH TEST OPERATOR AGENCY: DeSoto County SO
ARRESTING OFFICER: MILLER, PAUL
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 02/23/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:23

SUBJECT'S LAST NAME: RINCON

DATE/TIME OF TEST: 03/10/2017 02:52:17

Diagnostics Check	OK	02:43
Air Blank	0.000	02:43
Control Test	0.078	02:44
Air Blank	0.000	02:44
Subject Sample # 1	0.242	02:45
Breath Volume	1.230	02:45
Air Blank	0.000	02:46
Air Blank	0.000	02:48
Subject Sample # 2	0.236	02:50
Breath Volume	1.996	02:50
Air Blank	0.000	02:51
Control Test	0.079	02:51
Air Blank	0.000	02:52
Diagnostics Check	OK	02:52

Data Download Date/Time: 03/23/2017 09:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001341

INSTRUMENT REGISTERED TO: DESOTO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARROLL,STEVEN,C
BREATH TEST OPERATOR AGENCY: DeSoto County SO
ARRESTING OFFICER: MCLAUGHLIN,TERENCE
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 02/23/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: BAKER

DATE/TIME OF TEST: 03/17/2017 02:43:57

Diagnostics Check	OK	02:37
Air Blank	0.000	02:38
Control Test	0.080	02:38
Air Blank	0.000	02:39
Subject Sample # 1	0.151	02:39
Breath Volume	3.132	02:39
Air Blank	0.000	02:40
Air Blank	0.000	02:42
Subject Sample # 2	0.149	02:42
Breath Volume	3.621	02:42
Air Blank	0.000	02:43
Control Test	0.079	02:43
Air Blank	0.000	02:43
Diagnostics Check	OK	02:43

Data Download Date/Time: 03/23/2017 09:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001341

INSTRUMENT REGISTERED TO: DESOTO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARROLL,STEVEN,C
BREATH TEST OPERATOR AGENCY: DeSoto County SO
ARRESTING OFFICER: EVANS,GARY
ARRESTING OFFICER AGENCY: Arcadia PD
LAST AGENCY INSPECTION: 02/23/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: OTHER
ARREST TIME: 19:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:15

SUBJECT'S LAST NAME: MUNDO

DATE/TIME OF TEST: 03/21/2017 21:45:39

Diagnostics Check	OK	21:39
Air Blank	0.000	21:40
Control Test	0.079	21:40
Air Blank	0.000	21:40
Subject Sample # 1	0.000	21:41
Breath Volume	3.039	21:41
Air Blank	0.000	21:41
Air Blank	0.000	21:43
Subject Sample # 2	0.000	21:44
Breath Volume	2.855	21:44
Air Blank	0.000	21:44
Control Test	0.080	21:45
Air Blank	0.000	21:45
Diagnostics Check	OK	21:45

Data Download Date/Time: 03/23/2017 09:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

INSTRUMENT REGISTERED TO: EUSTIS P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STROUP,STEVE,A
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: GRIEVES,MATT
ARRESTING OFFICER AGENCY: Eustis PD
LAST AGENCY INSPECTION: 02/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 00:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:34

SUBJECT'S LAST NAME: MULLEN

DATE/TIME OF TEST: 03/03/2017 02:02:01

Diagnostics Check	OK	01:59	
Air Blank	0.000	01:59	
Control Test	0.079	01:59	
Air Blank	0.000	02:00	
Subject Sample	# 1	02:00	Subject Test Refused
Breath Volume	0.000	02:00	
Air Blank	0.000	02:01	
Control Test	0.079	02:01	
Air Blank	0.000	02:01	
Diagnostics Check	OK	02:02	

Data Download Date/Time: 03/27/2017 18:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

INSTRUMENT REGISTERED TO: EUSTIS P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HACKETT,SEAN,J

BREATH TEST OPERATOR AGENCY: Eustis PD

ARRESTING OFFICER: GRIEVES,MATTHEW

ARRESTING OFFICER AGENCY: Eustis PD

LAST AGENCY INSPECTION: 02/24/2017

Date or Time Changed: No

DRY GAS STANDARD LOT #: 203498

DRY GAS EXPIRY DATE: 12/10/2018

VIOLATION CODE: DUI

ARREST TIME: 23:09

GENDER: Male

OBSERVATION PERIOD BEGAN: 00:04

SUBJECT'S LAST NAME: DEJARNETTE

DATE/TIME OF TEST: 03/04/2017 00:36:59

Diagnostics Check	OK	00:30
Air Blank	0.000	00:31
Control Test	0.080	00:31
Air Blank	0.000	00:32
Subject Sample # 1	0.000	00:32
Breath Volume	3.378	00:32
Air Blank	0.000	00:33
Air Blank	0.000	00:35
Subject Sample # 2	0.000	00:35
Breath Volume	4.828	00:35
Air Blank	0.000	00:36
Control Test	0.078	00:36
Air Blank	0.000	00:36
Diagnostics Check	OK	00:36

Data Download Date/Time: 03/27/2017 18:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

INSTRUMENT REGISTERED TO: EUSTIS P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCGEE, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: LEDLOW, AARON
ARRESTING OFFICER AGENCY: Eustis PD
LAST AGENCY INSPECTION: 02/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 04:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: OLIVE

DATE/TIME OF TEST: 03/11/2017 04:56:42

Diagnostics Check	OK	04:46
Air Blank	0.000	04:46
Control Test	0.079	04:47
Air Blank	0.000	04:47
Subject Sample # 1	0.197	04:48
Breath Volume	1.742	04:48
Air Blank	0.000	04:49
Air Blank	0.000	04:51
Subject Sample # 2	0.229	04:51
Breath Volume	2.835	04:51
Air Blank	0.000	04:52
Air Blank	0.000	04:54
Subject Sample # 3	0.208	04:55
Breath Volume	3.617	04:55
Air Blank	0.000	04:55
Control Test	0.079	04:56
Air Blank	0.000	04:56
Diagnostics Check	OK	04:56

Data Download Date/Time: 03/27/2017 18:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

INSTRUMENT REGISTERED TO: EUSTIS P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCGEE, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: PERRY, WAYNE
ARRESTING OFFICER AGENCY: Eustis PD
LAST AGENCY INSPECTION: 02/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 15:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:08

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 03/24/2017 17:03:02

Diagnostics Check	OK	16:51	
Air Blank	0.000	16:51	
Control Test	0.079	16:52	
Air Blank	0.000	16:52	
Subject Sample	# 1 0.176	16:55	Volume Not Met
Breath Volume	0.218	16:55	
Air Blank	0.000	16:56	
Air Blank	0.000	16:58	
Subject Sample	# 2 0.212	17:01	Volume Not Met
Breath Volume	0.105	17:01	
Air Blank	0.000	17:02	
Control Test	0.078	17:02	
Air Blank	0.000	17:02	
Diagnostics Check	OK	17:03	

Data Download Date/Time: 03/27/2017 18:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

INSTRUMENT REGISTERED TO: EUSTIS P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCGEE, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: PERRY, WAYNE
ARRESTING OFFICER AGENCY: Eustis PD
LAST AGENCY INSPECTION: 02/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 15:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:08

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 03/24/2017 17:42:48

Diagnostics Check	OK	17:29	
Air Blank	0.000	17:30	
Control Test	0.079	17:30	
Air Blank	0.000	17:31	
Subject Sample	# 1 0.164	17:34	Slope Not Level
Breath Volume	1.195	17:34	
Air Blank	0.000	17:34	
Air Blank	0.000	17:36	
Subject Sample	# 2 0.204	17:37	
Breath Volume	2.468	17:37	
Air Blank	0.000	17:38	
Air Blank	0.000	17:39	
Subject Sample	# 3 0.197	17:41	
Breath Volume	2.554	17:41	
Air Blank	0.000	17:41	
Control Test	0.078	17:42	
Air Blank	0.000	17:42	
Diagnostics Check	OK	17:42	

Data Download Date/Time: 03/27/2017 18:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

INSTRUMENT REGISTERED TO: EUSTIS P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: STROUP,STEVE,A

BREATH TEST OPERATOR AGENCY: Eustis PD

ARRESTING OFFICER: LEDLOW,AARON

ARRESTING OFFICER AGENCY: Eustis PD

LAST AGENCY INSPECTION: 02/24/2017

Date or Time Changed: No

DRY GAS STANDARD LOT #: 203498

DRY GAS EXPIRY DATE: 12/10/2018

VIOLATION CODE: DUI

ARREST TIME: 21:56

GENDER: Female

OBSERVATION PERIOD BEGAN: 22:13

SUBJECT'S LAST NAME: EATON

DATE/TIME OF TEST: 03/25/2017 22:46:42

Diagnostics Check	OK	22:42	
Air Blank	0.000	22:43	
Control Test	0.079	22:43	
Air Blank	0.000	22:43	
Subject Sample	# 1	22:46	Sequence Aborted
Breath Volume	0.000	22:46	
Air Blank	0.000	22:46	

Data Download Date/Time: 03/27/2017 18:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

INSTRUMENT REGISTERED TO: EUSTIS P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: STROUP,STEVE,A

BREATH TEST OPERATOR AGENCY: Eustis PD

ARRESTING OFFICER: LEDLOW,AARON

ARRESTING OFFICER AGENCY: Eustis PD

LAST AGENCY INSPECTION: 02/24/2017

Date or Time Changed: No

DRY GAS STANDARD LOT #: 203498

DRY GAS EXPIRY DATE: 12/10/2018

VIOLATION CODE: DUI

ARREST TIME: 21:56

GENDER: Female

OBSERVATION PERIOD BEGAN: 22:13

SUBJECT'S LAST NAME: EATON

DATE/TIME OF TEST: 03/25/2017 23:00:40

Diagnostics Check	OK	22:48	
Air Blank	0.000	22:49	
Control Test	0.076	22:49	
Air Blank	0.000	22:50	
Subject Sample	# 1 0.206	22:53	Volume Not Met
Breath Volume	0.082	22:53	
Air Blank	0.000	22:54	
Air Blank	0.000	22:55	
Subject Sample	# 2	22:59	No Sample Provided
Breath Volume	0.000	22:59	
Air Blank	0.000	22:59	
Control Test	0.077	23:00	
Air Blank	0.000	23:00	
Diagnostics Check	OK	23:00	

Data Download Date/Time: 03/27/2017 18:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

INSTRUMENT REGISTERED TO: EUSTIS P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: STROUP,STEVE,A

BREATH TEST OPERATOR AGENCY: Eustis PD

ARRESTING OFFICER: LEDLOW,AARON

ARRESTING OFFICER AGENCY: Eustis PD

LAST AGENCY INSPECTION: 02/24/2017

Date or Time Changed: Yes

DRY GAS STANDARD LOT #: 203498

DRY GAS EXPIRY DATE: 12/10/2018

VIOLATION CODE: DUI

ARREST TIME: 21:56

GENDER: Female

OBSERVATION PERIOD BEGAN: 22:13

SUBJECT'S LAST NAME: EATON

DATE/TIME OF TEST: 03/25/2017 23:38:09

Diagnostics Check	OK	23:29	
Air Blank	0.000	23:30	
Control Test	0.078	23:30	
Air Blank	0.000	23:31	
Subject Sample	# 1 0.180	23:34	Volume Not Met
Breath Volume	1.082	23:34	
Air Blank	0.000	23:34	
Air Blank	0.000	23:36	
Subject Sample	# 2	23:36	Subject Test Refused
Breath Volume	0.000	23:36	
Air Blank	0.000	23:37	
Control Test	0.076	23:37	
Air Blank	0.000	23:38	
Diagnostics Check	OK	23:38	

Data Download Date/Time: 03/27/2017 18:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSTYAL,MATT,A
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: FOWLER,TYLER
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 02/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:53

SUBJECT'S LAST NAME: KIWCZAK

DATE/TIME OF TEST: 03/10/2017 03:19:44

Diagnostics Check	OK	03:13
Air Blank	0.000	03:13
Control Test	0.079	03:14
Air Blank	0.000	03:14
Subject Sample # 1	0.116	03:15
Breath Volume	2.125	03:15
Air Blank	0.000	03:15
Air Blank	0.000	03:17
Subject Sample # 2	0.112	03:18
Breath Volume	1.742	03:18
Air Blank	0.000	03:18
Control Test	0.079	03:19
Air Blank	0.000	03:19
Diagnostics Check	OK	03:19

Data Download Date/Time: 03/30/2017 08:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSTYAL,MATT,A
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: FOWLER,TYLER
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 02/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: PFOHLER

DATE/TIME OF TEST: 03/11/2017 01:37:02

Diagnostics Check	OK	01:30
Air Blank	0.000	01:31
Control Test	0.078	01:31
Air Blank	0.000	01:31
Subject Sample # 1	0.182	01:32
Breath Volume	3.210	01:32
Air Blank	0.000	01:32
Air Blank	0.000	01:34
Subject Sample # 2	0.171	01:35
Breath Volume	2.886	01:35
Air Blank	0.000	01:36
Control Test	0.078	01:36
Air Blank	0.000	01:36
Diagnostics Check	OK	01:37

Data Download Date/Time: 03/30/2017 08:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSTYAL,MATT,A
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: KOSTYAL,MATT
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 02/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: VISSER

DATE/TIME OF TEST: 03/11/2017 02:57:22

Diagnostics Check	OK	02:50
Air Blank	0.000	02:51
Control Test	0.079	02:51
Air Blank	0.000	02:52
Subject Sample # 1	0.074	02:52
Breath Volume	2.960	02:52
Air Blank	0.000	02:53
Air Blank	0.000	02:55
Subject Sample # 2	0.075	02:55
Breath Volume	2.886	02:55
Air Blank	0.000	02:56
Control Test	0.078	02:56
Air Blank	0.000	02:57
Diagnostics Check	OK	02:57

Data Download Date/Time: 03/30/2017 08:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PUFAHL,G,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: FOWLER,T
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 02/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:24

SUBJECT'S LAST NAME: MULCAHY

DATE/TIME OF TEST: 03/29/2017 23:51:21

Diagnostics Check	OK	23:45
Air Blank	0.000	23:45
Control Test	0.080	23:45
Air Blank	0.000	23:46
Subject Sample # 1	0.000	23:47
Breath Volume	3.164	23:47
Air Blank	0.000	23:47
Air Blank	0.000	23:49
Subject Sample # 2	0.000	23:49
Breath Volume	2.996	23:49
Air Blank	0.000	23:50
Control Test	0.079	23:50
Air Blank	0.000	23:51
Diagnostics Check	OK	23:51

Data Download Date/Time: 03/30/2017 08:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSTYAL,MATT,A
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: KOSTYAL,MATT
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 02/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A2
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:06

SUBJECT'S LAST NAME: RIDGEWAY

DATE/TIME OF TEST: 03/01/2017 03:38:05

Diagnostics Check	OK	03:30
Air Blank	0.000	03:31
Control Test	0.077	03:31
Air Blank	0.000	03:32
Subject Sample # 1	0.097	03:33
Breath Volume	1.191	03:33
Air Blank	0.000	03:33
Air Blank	0.000	03:35
Subject Sample # 2	0.096	03:36
Breath Volume	1.417	03:36
Air Blank	0.000	03:37
Control Test	0.078	03:37
Air Blank	0.000	03:38
Diagnostics Check	OK	03:38

Data Download Date/Time: 03/30/2017 07:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSTYAL,MATT,A
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: KOSTYAL,MATT
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 02/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A2
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: DE LIRA ZAVALA

DATE/TIME OF TEST: 03/02/2017 02:30:28

Diagnostics Check	OK	02:23
Air Blank	0.000	02:24
Control Test	0.077	02:24
Air Blank	0.000	02:24
Subject Sample # 1	0.197	02:25
Breath Volume	1.367	02:25
Air Blank	0.000	02:26
Air Blank	0.000	02:28
Subject Sample # 2	0.193	02:29
Breath Volume	1.207	02:29
Air Blank	0.000	02:29
Control Test	0.078	02:29
Air Blank	0.000	02:30
Diagnostics Check	OK	02:30

Data Download Date/Time: 03/30/2017 07:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSS,MICHAEL,K
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: PERKINS,OFFICER
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 02/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A2
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 15:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:10

SUBJECT'S LAST NAME: HURST

DATE/TIME OF TEST: 03/02/2017 16:39:31

Diagnostics Check	OK	16:33
Air Blank	0.000	16:34
Control Test	0.079	16:34
Air Blank	0.000	16:34
Subject Sample # 1	0.015	16:35
Breath Volume	3.093	16:35
Air Blank	0.000	16:35
Air Blank	0.000	16:37
Subject Sample # 2	0.015	16:38
Breath Volume	3.445	16:38
Air Blank	0.000	16:38
Control Test	0.079	16:38
Air Blank	0.000	16:39
Diagnostics Check	OK	16:39

Data Download Date/Time: 03/30/2017 07:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COSS,MICHAEL,G
BREATH TEST OPERATOR AGENCY: North Port PD
ARRESTING OFFICER: NICK,AARON
ARRESTING OFFICER AGENCY: North Port PD
LAST AGENCY INSPECTION: 02/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A2
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:13

SUBJECT'S LAST NAME: TIRPAK

DATE/TIME OF TEST: 03/02/2017 22:40:54

Diagnostics Check	OK	22:35
Air Blank	0.000	22:35
Control Test	0.079	22:35
Air Blank	0.000	22:36
Subject Sample # 1	0.000	22:36
Breath Volume	1.843	22:36
Air Blank	0.000	22:37
Air Blank	0.000	22:39
Subject Sample # 2	0.000	22:39
Breath Volume	1.367	22:39
Air Blank	0.000	22:40
Control Test	0.079	22:40
Air Blank	0.000	22:40
Diagnostics Check	OK	22:40

Data Download Date/Time: 03/30/2017 07:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSTYAL,MATT,A
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: KOSTYAL,MATT
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 02/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A2
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: MARBLE

DATE/TIME OF TEST: 03/03/2017 01:45:38

Diagnostics Check	OK	01:38
Air Blank	0.000	01:38
Control Test	0.079	01:39
Air Blank	0.000	01:39
Subject Sample # 1	0.137	01:40
Breath Volume	1.878	01:40
Air Blank	0.000	01:40
Air Blank	0.000	01:42
Subject Sample # 2	0.135	01:44
Breath Volume	1.484	01:44
Air Blank	0.000	01:44
Control Test	0.079	01:45
Air Blank	0.000	01:45
Diagnostics Check	OK	01:45

Data Download Date/Time: 03/30/2017 07:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLINT,SERGEANT,C
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: FLINT,SERGEANT
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 02/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A2
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 17:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:35

SUBJECT'S LAST NAME: FURTADO

DATE/TIME OF TEST: 03/03/2017 19:03:12

Diagnostics Check	OK	18:56
Air Blank	0.000	18:56
Control Test	0.079	18:56
Air Blank	0.000	18:57
Subject Sample # 1	0.198	18:58
Breath Volume	1.472	18:58
Air Blank	0.000	18:58
Air Blank	0.000	19:00
Subject Sample # 2	0.184	19:01
Breath Volume	1.101	19:01
Air Blank	0.000	19:02
Control Test	0.078	19:02
Air Blank	0.000	19:03
Diagnostics Check	OK	19:03

Data Download Date/Time: 03/30/2017 07:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,NICKKI,N
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: SCHUMANN,D
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 02/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A2
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:13

SUBJECT'S LAST NAME: CARTER

DATE/TIME OF TEST: 03/04/2017 21:50:42

Diagnostics Check	OK	21:44
Air Blank	0.000	21:44
Control Test	0.079	21:45
Air Blank	0.000	21:45
Subject Sample # 1	0.155	21:46
Breath Volume	2.296	21:46
Air Blank	0.000	21:46
Air Blank	0.000	21:48
Subject Sample # 2	0.154	21:49
Breath Volume	2.250	21:49
Air Blank	0.000	21:49
Control Test	0.079	21:50
Air Blank	0.000	21:50
Diagnostics Check	OK	21:50

Data Download Date/Time: 03/30/2017 07:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLINT,SERGEANT,C
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: FLINT,SERGEANT
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 02/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A2
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: RODRIGUEZ RODRIGUEZ

DATE/TIME OF TEST: 03/05/2017 02:23:01

Diagnostics Check	OK	02:16
Air Blank	0.000	02:16
Control Test	0.080	02:17
Air Blank	0.000	02:17
Subject Sample # 1	0.122	02:18
Breath Volume	1.187	02:18
Air Blank	0.000	02:19
Air Blank	0.000	02:20
Subject Sample # 2	0.133	02:21
Breath Volume	2.250	02:21
Air Blank	0.000	02:22
Control Test	0.079	02:22
Air Blank	0.000	02:22
Diagnostics Check	OK	02:23

Data Download Date/Time: 03/30/2017 07:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUBRANO,DEPUTY,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: SLOWICK,C
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 02/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A2
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 18:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:28

SUBJECT'S LAST NAME: BASSOE

DATE/TIME OF TEST: 03/05/2017 21:03:51

Diagnostics Check	OK	20:57
Air Blank	0.000	20:57
Control Test	0.079	20:58
Air Blank	0.000	20:58
Subject Sample # 1	0.077	20:59
Breath Volume	2.542	20:59
Air Blank	0.000	20:59
Air Blank	0.000	21:01
Subject Sample # 2	0.077	21:02
Breath Volume	2.519	21:02
Air Blank	0.000	21:02
Control Test	0.079	21:03
Air Blank	0.000	21:03
Diagnostics Check	OK	21:03

Data Download Date/Time: 03/30/2017 07:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSTYAL,MATT,A
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: KOSTYAL,MATT
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 02/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A2
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:46

SUBJECT'S LAST NAME: FANNING

DATE/TIME OF TEST: 03/10/2017 01:12:46

Diagnostics Check	OK	01:06
Air Blank	0.000	01:07
Control Test	0.079	01:07
Air Blank	0.000	01:07
Subject Sample # 1	0.151	01:08
Breath Volume	3.464	01:08
Air Blank	0.000	01:08
Air Blank	0.000	01:10
Subject Sample # 2	0.147	01:11
Breath Volume	2.558	01:11
Air Blank	0.000	01:11
Control Test	0.077	01:12
Air Blank	0.000	01:12
Diagnostics Check	OK	01:12

Data Download Date/Time: 03/30/2017 07:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANDERSON,DEPUTY,C
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BOBITZ,OFFICER
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 02/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A2
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: B535001978390

DATE/TIME OF TEST: 03/11/2017 01:13:56

Diagnostics Check	OK	01:05
Air Blank	0.000	01:05
Control Test	0.079	01:05
Air Blank	0.000	01:06
Subject Sample # 1	0.244	01:06
Breath Volume	1.738	01:06
Air Blank	0.000	01:07
Air Blank	0.000	01:09
Subject Sample # 2	0.212	01:09
Breath Volume	1.808	01:09
Air Blank	0.000	01:10
Air Blank	0.000	01:12
Subject Sample # 3	0.218	01:12
Breath Volume	2.453	01:12
Air Blank	0.000	01:13
Control Test	0.078	01:13
Air Blank	0.000	01:13
Diagnostics Check	OK	01:13

Data Download Date/Time: 03/30/2017 07:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARTHUR,MELVIN,J
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: ARTHUR,MELVIN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 02/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A2
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:23

SUBJECT'S LAST NAME: TREUBERT

DATE/TIME OF TEST: 03/11/2017 23:50:32

Diagnostics Check	OK	23:44
Air Blank	0.000	23:44
Control Test	0.079	23:44
Air Blank	0.000	23:45
Subject Sample # 1	0.197	23:46
Breath Volume	3.964	23:46
Air Blank	0.000	23:46
Air Blank	0.000	23:48
Subject Sample # 2	0.191	23:49
Breath Volume	3.031	23:49
Air Blank	0.000	23:49
Control Test	0.079	23:49
Air Blank	0.000	23:50
Diagnostics Check	OK	23:50

Data Download Date/Time: 03/30/2017 07:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GELLER,DEBORAH,A
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: DEFRANCISCO,OFFICER
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 02/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A2
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 10:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:52

SUBJECT'S LAST NAME: HUNTZINGER

DATE/TIME OF TEST: 03/14/2017 11:24:38

Diagnostics Check	OK	11:18
Air Blank	0.000	11:19
Control Test	0.080	11:19
Air Blank	0.000	11:19
Subject Sample # 1	0.000	11:20
Breath Volume	2.492	11:20
Air Blank	0.000	11:20
Air Blank	0.000	11:22
Subject Sample # 2	0.000	11:23
Breath Volume	2.652	11:23
Air Blank	0.000	11:23
Control Test	0.080	11:24
Air Blank	0.000	11:24
Diagnostics Check	OK	11:24

Data Download Date/Time: 03/30/2017 07:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALLACE,JENNIFER,L
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: WALLACE,JENNIFER
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 02/27/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 28515080A2
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:56

SUBJECT'S LAST NAME: RAMON

DATE/TIME OF TEST: 03/16/2017 02:24:21

Diagnostics Check	OK	02:17
Air Blank	0.000	02:18
Control Test	0.079	02:18
Air Blank	0.000	02:18
Subject Sample # 1	0.210	02:20
Breath Volume	1.937	02:20
Air Blank	0.000	02:20
Air Blank	0.000	02:22
Subject Sample # 2	0.204	02:22
Breath Volume	1.351	02:22
Air Blank	0.000	02:23
Control Test	0.079	02:23
Air Blank	0.000	02:24
Diagnostics Check	OK	02:24

Data Download Date/Time: 03/30/2017 07:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLYNN,MICHAEL,T
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: SCHUMANN,D
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 02/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A2
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:29

SUBJECT'S LAST NAME: REHAK

DATE/TIME OF TEST: 03/16/2017 04:00:24

Diagnostics Check	OK	03:53
Air Blank	0.000	03:54
Control Test	0.080	03:54
Air Blank	0.000	03:54
Subject Sample # 1	0.172	03:55
Breath Volume	2.019	03:55
Air Blank	0.000	03:56
Air Blank	0.000	03:58
Subject Sample # 2	0.166	03:58
Breath Volume	1.214	03:58
Air Blank	0.000	03:59
Control Test	0.079	03:59
Air Blank	0.000	04:00
Diagnostics Check	OK	04:00

Data Download Date/Time: 03/30/2017 07:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANDERSON,DEPUTY,C
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: COLEMAN,OFFICER
ARRESTING OFFICER AGENCY: North Port PD
LAST AGENCY INSPECTION: 02/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A2
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:34

SUBJECT'S LAST NAME: LANE

DATE/TIME OF TEST: 03/16/2017 05:04:29

Diagnostics Check	OK	04:58
Air Blank	0.000	04:58
Control Test	0.079	04:58
Air Blank	0.000	04:59
Subject Sample # 1	0.062	05:00
Breath Volume	4.304	05:00
Air Blank	0.000	05:00
Air Blank	0.000	05:02
Subject Sample # 2	0.062	05:03
Breath Volume	3.796	05:03
Air Blank	0.000	05:03
Control Test	0.079	05:03
Air Blank	0.000	05:04
Diagnostics Check	OK	05:04

Data Download Date/Time: 03/30/2017 07:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HEYMAN,LUKE,S
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: HEYMAN,LUKE
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 02/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A2
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 09:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:35

SUBJECT'S LAST NAME: KOCH

DATE/TIME OF TEST: 03/16/2017 10:02:18

Diagnostics Check	OK	09:55
Air Blank	0.000	09:56
Control Test	0.080	09:56
Air Blank	0.000	09:56
Subject Sample # 1	0.029	09:57
Breath Volume	2.308	09:57
Air Blank	0.000	09:58
Air Blank	0.000	09:59
Subject Sample # 2	0.029	10:00
Breath Volume	1.343	10:00
Air Blank	0.000	10:01
Control Test	0.080	10:01
Air Blank	0.000	10:02
Diagnostics Check	OK	10:02

Data Download Date/Time: 03/30/2017 07:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSS,MICHAEL,K
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: PERKINS,OFFICER
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 02/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A2
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 11:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:30

SUBJECT'S LAST NAME: BLAIKIE

DATE/TIME OF TEST: 03/16/2017 12:25:28

Diagnostics Check	OK	12:18
Air Blank	0.000	12:19
Control Test	0.080	12:19
Air Blank	0.000	12:20
Subject Sample # 1	0.000	12:21
Breath Volume	1.878	12:21
Air Blank	0.000	12:21
Air Blank	0.000	12:23
Subject Sample # 2	0.000	12:24
Breath Volume	1.371	12:24
Air Blank	0.000	12:24
Control Test	0.080	12:24
Air Blank	0.000	12:25
Diagnostics Check	OK	12:25

Data Download Date/Time: 03/30/2017 07:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NICK,A.,
BREATH TEST OPERATOR AGENCY: North Port PD
ARRESTING OFFICER: NICK,A.
ARRESTING OFFICER AGENCY: North Port PD
LAST AGENCY INSPECTION: 02/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A2
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 17:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:25

SUBJECT'S LAST NAME: GRINER

DATE/TIME OF TEST: 03/16/2017 19:51:50

Diagnostics Check	OK	19:45
Air Blank	0.000	19:46
Control Test	0.080	19:46
Air Blank	0.000	19:47
Subject Sample # 1	0.043	19:47
Breath Volume	2.894	19:47
Air Blank	0.000	19:48
Air Blank	0.000	19:50
Subject Sample # 2	0.039	19:50
Breath Volume	2.949	19:50
Air Blank	0.000	19:50
Control Test	0.079	19:51
Air Blank	0.000	19:51
Diagnostics Check	OK	19:51

Data Download Date/Time: 03/30/2017 07:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NICK,A,
BREATH TEST OPERATOR AGENCY: North Port PD
ARRESTING OFFICER: NICK,A
ARRESTING OFFICER AGENCY: North Port PD
LAST AGENCY INSPECTION: 02/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A2
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: CUMBERLAND

DATE/TIME OF TEST: 03/17/2017 22:41:33

Diagnostics Check	OK	22:40	
Air Blank	0.000	22:40	
Control Test	0.079	22:40	
Air Blank		22:41	Improper Sample
Air Blank	0.000	22:41	

Data Download Date/Time: 03/30/2017 07:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NICK,A,
BREATH TEST OPERATOR AGENCY: North Port PD
ARRESTING OFFICER: NICK,A
ARRESTING OFFICER AGENCY: North Port PD
LAST AGENCY INSPECTION: 02/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A2
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: CUMBERALND

DATE/TIME OF TEST: 03/17/2017 22:49:56

Diagnostics Check	OK	22:43
Air Blank	0.000	22:43
Control Test	0.080	22:44
Air Blank	0.000	22:44
Subject Sample # 1	0.135	22:45
Breath Volume	2.804	22:45
Air Blank	0.000	22:46
Air Blank	0.000	22:47
Subject Sample # 2	0.128	22:48
Breath Volume	4.050	22:48
Air Blank	0.000	22:49
Control Test	0.079	22:49
Air Blank	0.000	22:49
Diagnostics Check	OK	22:49

Data Download Date/Time: 03/30/2017 07:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,NICKKI,N
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: PERRY,M
ARRESTING OFFICER AGENCY: Venice PD
LAST AGENCY INSPECTION: 02/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A2
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: WARING

DATE/TIME OF TEST: 03/18/2017 03:52:08

Diagnostics Check	OK	03:45
Air Blank	0.000	03:46
Control Test	0.080	03:46
Air Blank	0.000	03:47
Subject Sample # 1	0.200	03:47
Breath Volume	2.500	03:47
Air Blank	0.000	03:48
Air Blank	0.000	03:50
Subject Sample # 2	0.195	03:50
Breath Volume	2.281	03:50
Air Blank	0.000	03:51
Control Test	0.078	03:51
Air Blank	0.000	03:52
Diagnostics Check	OK	03:52

Data Download Date/Time: 03/30/2017 07:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUBRANO,DEPUTY,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: SLOWICK,C
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 02/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A2
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 18:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:50

SUBJECT'S LAST NAME: CRAWFORD

DATE/TIME OF TEST: 03/18/2017 20:23:15

Diagnostics Check	OK	20:17	
Air Blank	0.000	20:18	
Control Test	0.079	20:18	
Air Blank	0.000	20:18	
Subject Sample	# 1 0.217	20:19	
Breath Volume	1.914	20:19	
Air Blank	0.000	20:20	
Air Blank	0.000	20:22	
Subject Sample	# 2 0.220	20:22	
Breath Volume	1.699	20:22	
Air Blank		20:22	Improper Sample
Air Blank	0.000	20:23	

Data Download Date/Time: 03/30/2017 07:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VILLOTTI,JOHN,J
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: HYDE,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 02/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A2
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:57

SUBJECT'S LAST NAME: ENGLISH

DATE/TIME OF TEST: 03/20/2017 22:30:52

Diagnostics Check	OK	22:23
Air Blank	0.000	22:24
Control Test	0.079	22:24
Air Blank	0.000	22:24
Subject Sample # 1	0.168	22:26
Breath Volume	1.980	22:26
Air Blank	0.000	22:27
Air Blank	0.000	22:28
Subject Sample # 2	0.184	22:29
Breath Volume	2.210	22:29
Air Blank	0.000	22:29
Control Test	0.079	22:30
Air Blank	0.000	22:30
Diagnostics Check	OK	22:30

Data Download Date/Time: 03/30/2017 07:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWARD, MICHAEL, E
BREATH TEST OPERATOR AGENCY: North Port PD
ARRESTING OFFICER: BOETGER, ELI
ARRESTING OFFICER AGENCY: North Port PD
LAST AGENCY INSPECTION: 02/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A2
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 18:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:30

SUBJECT'S LAST NAME: HOWARD

DATE/TIME OF TEST: 03/21/2017 20:05:46

Diagnostics Check	OK	19:56
Air Blank	0.000	19:57
Control Test	0.079	19:57
Air Blank	0.000	19:58
Subject Sample # 1	0.297	19:58
Breath Volume	2.648	19:58
Air Blank	0.000	19:59
Air Blank	0.000	20:01
Subject Sample # 2	0.274	20:01
Breath Volume	1.695	20:01
Air Blank	0.000	20:02
Air Blank	0.000	20:03
Subject Sample # 3	0.271	20:04
Breath Volume	1.285	20:04
Air Blank	0.000	20:04
Control Test	0.078	20:05
Air Blank	0.000	20:05
Diagnostics Check	OK	20:05

Data Download Date/Time: 03/30/2017 07:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,DANIEL,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: LONG,OFFICER
ARRESTING OFFICER AGENCY: North Port PD
LAST AGENCY INSPECTION: 02/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A2
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:10

SUBJECT'S LAST NAME: CURTISS

DATE/TIME OF TEST: 03/22/2017 23:12:38

Diagnostics Check	OK	23:06
Air Blank	0.000	23:06
Control Test	0.079	23:06
Air Blank	0.000	23:07
Subject Sample # 1	0.116	23:08
Breath Volume	2.078	23:08
Air Blank	0.000	23:08
Air Blank	0.000	23:10
Subject Sample # 2	0.114	23:11
Breath Volume	1.246	23:11
Air Blank	0.000	23:11
Control Test	0.078	23:12
Air Blank	0.000	23:12
Diagnostics Check	OK	23:12

Data Download Date/Time: 03/30/2017 07:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SOBAS,PATRICIA,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: CASSELLI,R
ARRESTING OFFICER AGENCY: North Port PD
LAST AGENCY INSPECTION: 02/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A2
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:05

SUBJECT'S LAST NAME: FELIX

DATE/TIME OF TEST: 03/23/2017 05:34:58

Diagnostics Check	OK	05:27
Air Blank	0.000	05:27
Control Test	0.079	05:28
Air Blank	0.000	05:28
Subject Sample # 1	0.232	05:29
Breath Volume	1.636	05:29
Air Blank	0.000	05:30
Air Blank	0.000	05:32
Subject Sample # 2	0.239	05:33
Breath Volume	2.171	05:33
Air Blank	0.000	05:34
Control Test	0.079	05:34
Air Blank	0.000	05:34
Diagnostics Check	OK	05:34

Data Download Date/Time: 03/30/2017 07:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FOWLER, TYLRE, J
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: FOWLER, TYLER
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 02/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A2
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:57

SUBJECT'S LAST NAME: GOODE

DATE/TIME OF TEST: 03/23/2017 22:26:03

Diagnostics Check	OK	22:18
Air Blank	0.000	22:18
Control Test	0.078	22:19
Air Blank	0.000	22:19
Subject Sample # 1	0.212	22:21
Breath Volume	1.714	22:21
Air Blank	0.000	22:22
Air Blank	0.000	22:23
Subject Sample # 2	0.209	22:24
Breath Volume	1.558	22:24
Air Blank	0.000	22:25
Control Test	0.077	22:25
Air Blank	0.000	22:25
Diagnostics Check	OK	22:26

Data Download Date/Time: 03/30/2017 07:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MINGS,DEP G,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: MINGS,DEP G
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 02/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A2
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:57

SUBJECT'S LAST NAME: LUNNEY

DATE/TIME OF TEST: 03/24/2017 23:25:42

Diagnostics Check	OK	23:19
Air Blank	0.000	23:19
Control Test	0.079	23:19
Air Blank	0.000	23:20
Subject Sample # 1	0.153	23:21
Breath Volume	1.984	23:21
Air Blank	0.000	23:21
Air Blank	0.000	23:23
Subject Sample # 2	0.142	23:24
Breath Volume	1.519	23:24
Air Blank	0.000	23:24
Control Test	0.079	23:25
Air Blank	0.000	23:25
Diagnostics Check	OK	23:25

Data Download Date/Time: 03/30/2017 07:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURLEY,SHAUN,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: SCHUMANN,DANNY
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 02/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A2
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:46

SUBJECT'S LAST NAME: ALLEN

DATE/TIME OF TEST: 03/25/2017 01:43:50

Diagnostics Check	OK	01:37
Air Blank	0.000	01:37
Control Test	0.079	01:38
Air Blank	0.000	01:38
Subject Sample # 1	0.166	01:39
Breath Volume	2.546	01:39
Air Blank	0.000	01:39
Air Blank	0.000	01:41
Subject Sample # 2	0.166	01:42
Breath Volume	1.710	01:42
Air Blank	0.000	01:42
Control Test	0.079	01:43
Air Blank	0.000	01:43
Diagnostics Check	OK	01:43

Data Download Date/Time: 03/30/2017 07:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURLEY,SHAUN,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BISHOP,C
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 02/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A2
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:49

SUBJECT'S LAST NAME: DRESSIG

DATE/TIME OF TEST: 03/25/2017 23:21:31

Diagnostics Check	OK	23:12
Air Blank	0.000	23:12
Control Test	0.080	23:13
Air Blank	0.000	23:13
Subject Sample # 1	0.208	23:14
Breath Volume	1.257	23:14
Air Blank	0.000	23:14
Air Blank	0.000	23:16
Subject Sample # 2	0.231	23:17
Breath Volume	1.207	23:17
Air Blank	0.000	23:17
Air Blank	0.000	23:19
Subject Sample # 3	0.229	23:20
Breath Volume	2.312	23:20
Air Blank	0.000	23:20
Control Test	0.078	23:20
Air Blank	0.000	23:21
Diagnostics Check	OK	23:21

Data Download Date/Time: 03/30/2017 07:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DARLING,JOHN,P
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: SHOPE,K
ARRESTING OFFICER AGENCY: Venice PD
LAST AGENCY INSPECTION: 02/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A2
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 05:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:04

SUBJECT'S LAST NAME: LAUER

DATE/TIME OF TEST: 03/26/2017 07:53:36

Diagnostics Check	OK	07:46
Air Blank	0.000	07:47
Control Test	0.079	07:47
Air Blank	0.000	07:48
Subject Sample # 1	0.213	07:48
Breath Volume	1.574	07:48
Air Blank	0.000	07:49
Air Blank	0.000	07:51
Subject Sample # 2	0.207	07:52
Breath Volume	1.992	07:52
Air Blank	0.000	07:52
Control Test	0.078	07:53
Air Blank	0.000	07:53
Diagnostics Check	OK	07:53

Data Download Date/Time: 03/30/2017 07:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VILLOTTI,DEP,J
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: VILLOTTI,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 02/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A2
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:12

SUBJECT'S LAST NAME: FREY

DATE/TIME OF TEST: 03/26/2017 21:39:25

Diagnostics Check	OK	21:33
Air Blank	0.000	21:33
Control Test	0.079	21:33
Air Blank	0.000	21:34
Subject Sample # 1	0.154	21:34
Breath Volume	1.339	21:34
Air Blank	0.000	21:35
Air Blank	0.000	21:37
Subject Sample # 2	0.141	21:37
Breath Volume	1.382	21:37
Air Blank	0.000	21:38
Control Test	0.078	21:38
Air Blank	0.000	21:39
Diagnostics Check	OK	21:39

Data Download Date/Time: 03/30/2017 07:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FOWLER, TYLER, J
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: FOWLER, TYLER
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 03/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A2
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:59

SUBJECT'S LAST NAME: FERRAO

DATE/TIME OF TEST: 03/31/2017 22:25:56

Diagnostics Check	OK	22:19
Air Blank	0.000	22:20
Control Test	0.078	22:20
Air Blank	0.000	22:20
Subject Sample # 1	0.212	22:21
Breath Volume	2.562	22:21
Air Blank	0.000	22:22
Air Blank	0.000	22:24
Subject Sample # 2	0.212	22:24
Breath Volume	2.371	22:24
Air Blank	0.000	22:25
Control Test	0.078	22:25
Air Blank	0.000	22:25
Diagnostics Check	OK	22:25

Data Download Date/Time: 04/20/2017 11:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NICK,A,
BREATH TEST OPERATOR AGENCY: North Port PD
ARRESTING OFFICER: NICK,A
ARRESTING OFFICER AGENCY: North Port PD
LAST AGENCY INSPECTION: 03/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A2
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: FINOL

DATE/TIME OF TEST: 03/31/2017 23:09:53

Diagnostics Check	OK	23:00
Air Blank	0.000	23:01
Control Test	0.079	23:01
Air Blank	0.000	23:02
Subject Sample # 1	0.162	23:02
Breath Volume	4.281	23:02
Air Blank	0.000	23:03
Air Blank	0.000	23:05
Subject Sample # 2	0.139	23:05
Breath Volume	4.750	23:05
Air Blank	0.000	23:06
Air Blank	0.000	23:07
Subject Sample # 3	0.142	23:08
Breath Volume	3.843	23:08
Air Blank	0.000	23:08
Control Test	0.078	23:09
Air Blank	0.000	23:09
Diagnostics Check	OK	23:09

Data Download Date/Time: 04/20/2017 11:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SOWARDS,JACOB,T
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: GUYTON,JOHN
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 02/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 03/04/2017 04:38:32

Diagnostics Check	OK	04:24	
Air Blank	0.000	04:25	
Control Test	0.079	04:25	
Air Blank	0.000	04:26	
Subject Sample	# 1 0.138	04:27	
Breath Volume	1.101	04:27	
Air Blank	0.000	04:28	
Air Blank	0.000	04:30	
Subject Sample	# 2 0.130	04:33	Volume Not Met
Breath Volume	0.015	04:33	
Air Blank	0.000	04:34	
Air Blank	0.000	04:35	
Subject Sample	# 3 0.144	04:36	
Breath Volume	1.265	04:36	
Air Blank	0.000	04:37	
Control Test	0.077	04:37	
Air Blank	0.000	04:38	
Diagnostics Check	OK	04:38	

Data Download Date/Time: 03/22/2017 20:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEMEN,AARON,M
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: STEMEN,AARON
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 02/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:21

SUBJECT'S LAST NAME: GARGANO

DATE/TIME OF TEST: 03/11/2017 03:49:31

Diagnostics Check	OK	03:43
Air Blank	0.000	03:43
Control Test	0.079	03:44
Air Blank	0.000	03:44
Subject Sample # 1	0.245	03:45
Breath Volume	2.375	03:45
Air Blank	0.000	03:45
Air Blank	0.000	03:47
Subject Sample # 2	0.256	03:47
Breath Volume	2.421	03:47
Air Blank	0.000	03:48
Control Test	0.079	03:48
Air Blank	0.000	03:49
Diagnostics Check	OK	03:49

Data Download Date/Time: 03/22/2017 20:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEMEN,AARON,M
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: STEMEN,AARON
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 02/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 03/19/2017 02:20:52

Diagnostics Check	OK	02:14
Air Blank	0.000	02:15
Control Test	0.078	02:15
Air Blank	0.000	02:15
Subject Sample # 1	0.034	02:16
Breath Volume	1.910	02:16
Air Blank	0.000	02:17
Air Blank	0.000	02:18
Subject Sample # 2	0.033	02:19
Breath Volume	1.484	02:19
Air Blank	0.000	02:19
Control Test	0.078	02:20
Air Blank	0.000	02:20
Diagnostics Check	OK	02:20

Data Download Date/Time: 03/22/2017 20:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEMEN,AARON,M
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: STEMEN,AARON
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 03/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:36

SUBJECT'S LAST NAME: WORRICK

DATE/TIME OF TEST: 03/24/2017 03:07:32

Diagnostics Check	OK	02:59
Air Blank	0.000	03:00
Control Test	0.077	03:00
Air Blank	0.000	03:01
Subject Sample # 1	0.200	03:02
Breath Volume	1.585	03:02
Air Blank	0.000	03:03
Air Blank	0.000	03:05
Subject Sample # 2	0.202	03:06
Breath Volume	1.375	03:06
Air Blank	0.000	03:06
Control Test	0.078	03:06
Air Blank	0.000	03:07
Diagnostics Check	OK	03:07

Data Download Date/Time: 04/21/2017 21:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,THEODORE,D
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: JONES,THEODORE
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 03/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:52

SUBJECT'S LAST NAME: POLLOCK

DATE/TIME OF TEST: 03/25/2017 05:24:09

Diagnostics Check	OK	05:12	
Air Blank	0.000	05:13	
Control Test	0.078	05:13	
Air Blank	0.000	05:14	
Subject Sample	# 1	05:17	No Sample Provided
Breath Volume	0.000	05:17	
Air Blank	0.000	05:17	
Air Blank	0.000	05:19	
Subject Sample	# 2	05:22	No Sample Provided
Breath Volume	0.000	05:22	
Air Blank	0.000	05:23	
Control Test	0.078	05:23	
Air Blank	0.000	05:24	
Diagnostics Check	OK	05:24	

Data Download Date/Time: 04/21/2017 21:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SOWARDS,JACOB,T
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: GUYTON,JONATHAN
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 03/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:43

SUBJECT'S LAST NAME: MARTIN

DATE/TIME OF TEST: 03/27/2017 04:10:00

Diagnostics Check	OK	04:03
Air Blank	0.000	04:03
Control Test	0.078	04:04
Air Blank	0.000	04:04
Subject Sample # 1	0.000	04:05
Breath Volume	2.144	04:05
Air Blank	0.000	04:06
Air Blank	0.000	04:08
Subject Sample # 2	0.000	04:08
Breath Volume	2.476	04:08
Air Blank	0.000	04:09
Control Test	0.078	04:09
Air Blank	0.000	04:09
Diagnostics Check	OK	04:10

Data Download Date/Time: 04/21/2017 21:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEMEN,AARON,M
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: STEMEN,AARON
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 03/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:16

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 03/30/2017 00:44:44

Diagnostics Check	OK	00:36	
Air Blank	0.000	00:36	
Control Test	0.078	00:36	
Air Blank	0.000	00:37	
Subject Sample	# 1	00:40	No Sample Provided
Breath Volume	0.000	00:40	
Air Blank	0.000	00:41	
Air Blank	0.000	00:43	
Subject Sample	# 2	00:43	Subject Test Refused
Breath Volume	0.000	00:43	
Air Blank	0.000	00:43	
Control Test	0.079	00:44	
Air Blank	0.000	00:44	
Diagnostics Check	OK	00:44	

Data Download Date/Time: 04/21/2017 21:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001365

INSTRUMENT REGISTERED TO: LEVY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LESHER, TIFFANY, A
BREATH TEST OPERATOR AGENCY: Levy County SO
ARRESTING OFFICER: WILLIS, MORGAN
ARRESTING OFFICER AGENCY: FFWCC (Lake City)
LAST AGENCY INSPECTION: 02/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG619605
DRY GAS EXPIRY DATE: 07/14/2018
VIOLATION CODE: BUI
ARREST TIME: 17:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:24

SUBJECT'S LAST NAME: TRAVERS

DATE/TIME OF TEST: 03/02/2017 19:11:56

Diagnostics Check	OK	19:05
Air Blank	0.000	19:06
Control Test	0.081	19:06
Air Blank	0.000	19:07
Subject Sample # 1	0.000	19:07
Breath Volume	1.511	19:07
Air Blank	0.000	19:07
Air Blank	0.000	19:09
Subject Sample # 2	0.000	19:10
Breath Volume	1.460	19:10
Air Blank	0.000	19:11
Control Test	0.080	19:11
Air Blank	0.000	19:11
Diagnostics Check	OK	19:11

Data Download Date/Time: 03/13/2017 21:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001365

INSTRUMENT REGISTERED TO: LEVY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FULTON,DARIN,M
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: FULTON,DARIN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 02/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG619605
DRY GAS EXPIRY DATE: 07/14/2018
VIOLATION CODE: DUI
ARREST TIME: 09:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:30

SUBJECT'S LAST NAME: BODLE

DATE/TIME OF TEST: 03/03/2017 09:54:28

Diagnostics Check	OK	09:50	
Air Blank	0.000	09:50	
Control Test	0.081	09:51	
Air Blank	0.000	09:51	
Subject Sample	# 1	09:53	Subject Test Refused
Breath Volume	0.000	09:53	
Air Blank	0.000	09:53	
Control Test	0.081	09:53	
Air Blank	0.000	09:54	
Diagnostics Check	OK	09:54	

Data Download Date/Time: 03/13/2017 21:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001365

INSTRUMENT REGISTERED TO: LEVY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINKOFER,ZACHARY,R
BREATH TEST OPERATOR AGENCY: Levy County SO
ARRESTING OFFICER: ROE,J
ARRESTING OFFICER AGENCY: Levy County SO
LAST AGENCY INSPECTION: 02/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG619605
DRY GAS EXPIRY DATE: 07/14/2018
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:36

SUBJECT'S LAST NAME: BARKER

DATE/TIME OF TEST: 03/05/2017 03:08:57

Diagnostics Check	OK	03:06	
Air Blank	0.000	03:06	
Control Test	0.081	03:07	
Air Blank	0.000	03:07	
Subject Sample	# 1	03:07	Subject Test Refused
Breath Volume	0.000	03:07	
Air Blank		03:08	Ambient Fail
Air Blank		03:08	Purge Fail

Data Download Date/Time: 03/13/2017 21:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001365

INSTRUMENT REGISTERED TO: LEVY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANUS,GLEN,A
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: GANUS,GLEN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 02/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG619605
DRY GAS EXPIRY DATE: 07/14/2018
VIOLATION CODE: DUI
ARREST TIME: 21:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:19

SUBJECT'S LAST NAME: ROGERS

DATE/TIME OF TEST: 03/09/2017 22:42:38

Diagnostics Check	OK	22:39	
Air Blank	0.000	22:40	
Control Test	0.081	22:40	
Air Blank	0.000	22:41	
Subject Sample	# 1	22:41	Subject Test Refused
Breath Volume	0.000	22:41	
Air Blank	0.000	22:41	
Control Test	0.082	22:42	
Air Blank	0.000	22:42	
Diagnostics Check	OK	22:42	

Data Download Date/Time: 03/13/2017 21:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001365

INSTRUMENT REGISTERED TO: LEVY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNYDER,SCOTT,C
BREATH TEST OPERATOR AGENCY: Levy County SO
ARRESTING OFFICER: ADAMS,MATTHEW
ARRESTING OFFICER AGENCY: Levy County SO
LAST AGENCY INSPECTION: 02/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG619605
DRY GAS EXPIRY DATE: 07/14/2018
VIOLATION CODE: DUI
ARREST TIME: 03:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: RECTOR

DATE/TIME OF TEST: 03/11/2017 07:44:47

Diagnostics Check	OK	07:41	
Air Blank	0.000	07:42	
Control Test	0.081	07:42	
Air Blank	0.000	07:43	
Subject Sample	# 1	07:43	Subject Test Refused
Breath Volume	0.000	07:43	
Air Blank	0.000	07:43	
Control Test	0.082	07:44	
Air Blank	0.000	07:44	
Diagnostics Check	OK	07:44	

Data Download Date/Time: 03/13/2017 21:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001365

INSTRUMENT REGISTERED TO: LEVY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNYDER,SCOTT,C
BREATH TEST OPERATOR AGENCY: Levy County SO
ARRESTING OFFICER: VANDUREN,JUSTIN
ARRESTING OFFICER AGENCY: Levy County SO
LAST AGENCY INSPECTION: 02/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG619605
DRY GAS EXPIRY DATE: 07/14/2018
VIOLATION CODE: DUI
ARREST TIME: 05:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:50

SUBJECT'S LAST NAME: WITHROCK

DATE/TIME OF TEST: 03/12/2017 06:24:23

Diagnostics Check	OK	06:17
Air Blank	0.000	06:18
Control Test	0.081	06:18
Air Blank	0.000	06:19
Subject Sample # 1	0.131	06:19
Breath Volume	2.089	06:19
Air Blank	0.000	06:20
Air Blank	0.000	06:22
Subject Sample # 2	0.131	06:22
Breath Volume	1.980	06:22
Air Blank	0.000	06:23
Control Test	0.080	06:23
Air Blank	0.000	06:24
Diagnostics Check	OK	06:24

Data Download Date/Time: 03/13/2017 21:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001365

INSTRUMENT REGISTERED TO: LEVY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOWEN,MAYRA,I
BREATH TEST OPERATOR AGENCY: Levy County SO
ARRESTING OFFICER: SULLIVAN,R
ARRESTING OFFICER AGENCY: Levy County SO
LAST AGENCY INSPECTION: 03/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG619605
DRY GAS EXPIRY DATE: 07/14/2018
VIOLATION CODE: DUI
ARREST TIME: 21:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: BLACK

DATE/TIME OF TEST: 03/19/2017 22:15:47

Diagnostics Check	OK	22:08
Air Blank	0.000	22:08
Control Test	0.081	22:08
Air Blank	0.000	22:09
Subject Sample # 1	0.232	22:10
Breath Volume	3.375	22:10
Air Blank	0.000	22:11
Air Blank	0.000	22:12
Subject Sample # 2	0.221	22:13
Breath Volume	2.132	22:13
Air Blank	0.000	22:14
Control Test	0.080	22:15
Air Blank	0.000	22:15
Diagnostics Check	OK	22:15

Data Download Date/Time: 04/29/2017 20:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001365

INSTRUMENT REGISTERED TO: LEVY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINKOFER,ZACHARY,R
BREATH TEST OPERATOR AGENCY: Levy County SO
ARRESTING OFFICER: RIX,D
ARRESTING OFFICER AGENCY: Levy County SO
LAST AGENCY INSPECTION: 03/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG619605
DRY GAS EXPIRY DATE: 07/14/2018
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: FRAZIER

DATE/TIME OF TEST: 03/24/2017 02:58:21

Diagnostics Check	OK	02:51
Air Blank	0.000	02:51
Control Test	0.081	02:52
Air Blank	0.000	02:52
Subject Sample # 1	0.101	02:53
Breath Volume	1.566	02:53
Air Blank	0.000	02:54
Air Blank	0.000	02:56
Subject Sample # 2	0.104	02:56
Breath Volume	1.156	02:56
Air Blank	0.000	02:57
Control Test	0.080	02:57
Air Blank	0.000	02:58
Diagnostics Check	OK	02:58

Data Download Date/Time: 04/29/2017 20:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAIR, RONALD, J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ELLIS, ALEXANDER
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/03/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 00:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:22

SUBJECT'S LAST NAME: DUVERNAY

DATE/TIME OF TEST: 03/04/2017 00:51:21

Diagnostics Check	OK	00:42
Air Blank	0.000	00:43
Control Test	0.082	00:43
Air Blank	0.000	00:44
Subject Sample # 1	0.239	00:46
Breath Volume	1.320	00:46
Air Blank	0.000	00:47
Air Blank	0.000	00:48
Subject Sample # 2	0.231	00:49
Breath Volume	1.242	00:49
Air Blank	0.000	00:50
Control Test	0.080	00:50
Air Blank	0.000	00:51
Diagnostics Check	OK	00:51

Data Download Date/Time: 03/18/2017 21:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAIR, RONALD, J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WALDO, AARON
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 02/03/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 01:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:18

SUBJECT'S LAST NAME: FARLEY

DATE/TIME OF TEST: 03/04/2017 01:45:27

Diagnostics Check	OK	01:39
Air Blank	0.000	01:39
Control Test	0.082	01:39
Air Blank	0.000	01:40
Subject Sample # 1	0.106	01:40
Breath Volume	1.570	01:40
Air Blank	0.000	01:41
Air Blank	0.000	01:43
Subject Sample # 2	0.102	01:43
Breath Volume	1.929	01:43
Air Blank	0.000	01:44
Control Test	0.081	01:44
Air Blank	0.000	01:45
Diagnostics Check	OK	01:45

Data Download Date/Time: 03/18/2017 21:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAIR, RONALD, J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: CARTER, TIMOTHY
ARRESTING OFFICER AGENCY: Gulfport PD
LAST AGENCY INSPECTION: 02/03/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: UYEDA

DATE/TIME OF TEST: 03/04/2017 02:22:38

Diagnostics Check	OK	02:16
Air Blank	0.000	02:16
Control Test	0.082	02:17
Air Blank	0.000	02:17
Subject Sample # 1	0.037	02:18
Breath Volume	3.316	02:18
Air Blank	0.000	02:18
Air Blank	0.000	02:20
Subject Sample # 2	0.035	02:21
Breath Volume	2.222	02:21
Air Blank	0.000	02:21
Control Test	0.081	02:22
Air Blank	0.000	02:22
Diagnostics Check	OK	02:22

Data Download Date/Time: 03/18/2017 21:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAIR, RONALD, J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KARAYIANES, MICHAEL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 02/03/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 02:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:29

SUBJECT'S LAST NAME: TERRELL

DATE/TIME OF TEST: 03/04/2017 02:54:33

Diagnostics Check	OK	02:50	
Air Blank	0.000	02:50	
Control Test	0.082	02:50	
Air Blank	0.000	02:51	
Subject Sample	# 1	02:52	Subject Test Refused
Breath Volume	0.000	02:52	
Air Blank	0.000	02:53	
Control Test	0.082	02:53	
Air Blank	0.000	02:53	
Diagnostics Check	Fail	02:54	Temperature Regulation Test
Air Blank	0.000	02:54	

Data Download Date/Time: 03/18/2017 21:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001368

INSTRUMENT REGISTERED TO: NAS MAYPORT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OASIO,JOSE,A
BREATH TEST OPERATOR AGENCY: Mayport Naval Base
ARRESTING OFFICER: POLLOCK,DEREK
ARRESTING OFFICER AGENCY: Mayport Naval Base
LAST AGENCY INSPECTION: 02/14/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: PATTEN

DATE/TIME OF TEST: 03/02/2017 01:30:49

Diagnostics Check	OK	01:17	
Air Blank	0.000	01:17	
Control Test	0.078	01:17	
Air Blank	0.000	01:18	
Subject Sample	# 1 0.107	01:21	Volume Not Met
Breath Volume	0.000	01:21	
Air Blank	0.000	01:22	
Air Blank	0.000	01:23	
Subject Sample	# 2 0.148	01:25	
Breath Volume	1.820	01:25	
Air Blank	0.000	01:26	
Air Blank	0.000	01:27	
Subject Sample	# 3 0.141	01:29	
Breath Volume	1.750	01:29	
Air Blank	0.000	01:29	
Control Test	0.079	01:30	
Air Blank	0.000	01:30	
Diagnostics Check	OK	01:30	

Data Download Date/Time: 03/16/2017 08:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OKEEFE,ALEXIS,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: GALLUP,E
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 03/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: DAVIDSON

DATE/TIME OF TEST: 03/24/2017 02:54:16

Diagnostics Check	OK	02:47
Air Blank	0.000	02:48
Control Test	0.082	02:48
Air Blank	0.000	02:49
Subject Sample # 1	0.069	02:49
Breath Volume	1.722	02:49
Air Blank	0.000	02:50
Air Blank	0.000	02:51
Subject Sample # 2	0.063	02:52
Breath Volume	1.457	02:52
Air Blank	0.000	02:53
Control Test	0.082	02:53
Air Blank	0.000	02:54
Diagnostics Check	OK	02:54

Data Download Date/Time: 03/30/2017 15:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDEN, WILLIAM, L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ROBBINS, NATHAN
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 03/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: BEKKOUCHE

DATE/TIME OF TEST: 03/25/2017 01:37:11

Diagnostics Check	OK	01:31
Air Blank	0.000	01:31
Control Test	0.082	01:31
Air Blank	0.000	01:32
Subject Sample # 1	0.000	01:32
Breath Volume	2.820	01:32
Air Blank	0.000	01:33
Air Blank	0.000	01:35
Subject Sample # 2	0.000	01:35
Breath Volume	3.039	01:35
Air Blank	0.000	01:36
Control Test	0.082	01:36
Air Blank	0.000	01:37
Diagnostics Check	OK	01:37

Data Download Date/Time: 03/30/2017 15:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDEN, WILLIAM, L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DODGE-MARRA, KAITLYN
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 03/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: BONILLA-MEJIA

DATE/TIME OF TEST: 03/25/2017 04:34:23

Diagnostics Check	OK	04:27
Air Blank	0.000	04:27
Control Test	0.082	04:28
Air Blank	0.000	04:28
Subject Sample # 1	0.192	04:29
Breath Volume	3.125	04:29
Air Blank	0.000	04:30
Air Blank	0.000	04:32
Subject Sample # 2	0.189	04:32
Breath Volume	2.089	04:32
Air Blank	0.000	04:33
Control Test	0.081	04:33
Air Blank	0.000	04:34
Diagnostics Check	OK	04:34

Data Download Date/Time: 03/30/2017 15:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDEN, WILLIAM, L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MURPHY, SEAN
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 03/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: TOMLIN

DATE/TIME OF TEST: 03/26/2017 01:02:44

Diagnostics Check	OK	00:57	
Air Blank	0.000	00:57	
Control Test	0.082	00:58	
Air Blank	0.000	00:58	
Subject Sample	# 1	01:01	Subject Test Refused
Breath Volume	0.000	01:01	
Air Blank	0.000	01:01	
Control Test	0.083	01:02	
Air Blank	0.000	01:02	
Diagnostics Check	OK	01:02	

Data Download Date/Time: 03/30/2017 15:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OKEEFE,ALEXIS,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: NUTTER,M
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 03/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: ROWLEY

DATE/TIME OF TEST: 03/26/2017 03:59:44

Diagnostics Check	OK	03:53
Air Blank	0.000	03:53
Control Test	0.082	03:54
Air Blank	0.000	03:54
Subject Sample # 1	0.162	03:54
Breath Volume	3.144	03:54
Air Blank	0.000	03:55
Air Blank	0.000	03:57
Subject Sample # 2	0.149	03:58
Breath Volume	3.613	03:58
Air Blank	0.000	03:58
Control Test	0.081	03:59
Air Blank	0.000	03:59
Diagnostics Check	OK	03:59

Data Download Date/Time: 03/30/2017 15:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDEN, WILLIAM, L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ROBBINS, NATHAN
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 03/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: CAREY

DATE/TIME OF TEST: 03/26/2017 22:26:51

Diagnostics Check	OK	22:20
Air Blank	0.000	22:20
Control Test	0.083	22:20
Air Blank	0.000	22:21
Subject Sample # 1	0.176	22:22
Breath Volume	2.535	22:22
Air Blank	0.000	22:22
Air Blank	0.000	22:24
Subject Sample # 2	0.161	22:25
Breath Volume	2.917	22:25
Air Blank	0.000	22:25
Control Test	0.082	22:26
Air Blank	0.000	22:26
Diagnostics Check	OK	22:26

Data Download Date/Time: 03/30/2017 15:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSON,MATTHEW,P
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: CAMPBELL,T
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 02/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:12

SUBJECT'S LAST NAME: COLLINS

DATE/TIME OF TEST: 03/02/2017 00:42:37

Diagnostics Check	OK	00:36
Air Blank	0.000	00:36
Control Test	0.081	00:37
Air Blank	0.000	00:37
Subject Sample # 1	0.139	00:38
Breath Volume	2.128	00:38
Air Blank	0.000	00:38
Air Blank	0.000	00:40
Subject Sample # 2	0.145	00:41
Breath Volume	2.410	00:41
Air Blank	0.000	00:41
Control Test	0.080	00:42
Air Blank	0.000	00:42
Diagnostics Check	OK	00:42

Data Download Date/Time: 03/18/2017 11:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALTSTATT,MATTHEW,B
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: PALMER,Z
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 02/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 19:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:04

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 03/08/2017 20:32:37

Diagnostics Check	OK	20:24
Air Blank	0.000	20:25
Control Test	0.081	20:25
Air Blank	0.000	20:25
Subject Sample # 1	0.194	20:27
Breath Volume	1.945	20:27
Air Blank	0.000	20:28
Air Blank	0.000	20:30
Subject Sample # 2	0.188	20:31
Breath Volume	1.917	20:31
Air Blank	0.000	20:31
Control Test	0.080	20:32
Air Blank	0.000	20:32
Diagnostics Check	OK	20:32

Data Download Date/Time: 03/18/2017 11:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALTSTATT,MATTHEW,B
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: KOEHLER,A
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 02/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 19:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:38

SUBJECT'S LAST NAME: TACHELL

DATE/TIME OF TEST: 03/09/2017 20:07:26

Diagnostics Check	OK	19:58	
Air Blank	0.000	19:59	
Control Test	0.081	19:59	
Air Blank	0.000	20:00	
Subject Sample	# 1	20:03	No Sample Provided
Breath Volume	0.000	20:03	
Air Blank	0.000	20:03	
Air Blank	0.000	20:05	
Subject Sample	# 2	20:06	Subject Test Refused
Breath Volume	0.000	20:06	
Air Blank	0.000	20:06	
Control Test	0.081	20:06	
Air Blank	0.000	20:07	
Diagnostics Check	OK	20:07	

Data Download Date/Time: 03/18/2017 11:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALTSTATT,MATTHEW,B
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: PERKINS,T
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 02/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: WESTMORELAND

DATE/TIME OF TEST: 03/18/2017 01:36:55

Diagnostics Check	OK	01:30
Air Blank	0.000	01:30
Control Test	0.081	01:31
Air Blank	0.000	01:31
Subject Sample # 1	0.037	01:32
Breath Volume	3.890	01:32
Air Blank	0.000	01:32
Air Blank	0.000	01:34
Subject Sample # 2	0.037	01:35
Breath Volume	4.101	01:35
Air Blank	0.000	01:36
Control Test	0.081	01:36
Air Blank	0.000	01:36
Diagnostics Check	OK	01:36

Data Download Date/Time: 03/18/2017 11:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZIER,TAMMY,L
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: RULON,NICHOLAS
ARRESTING OFFICER AGENCY: Green Cove Springs
LAST AGENCY INSPECTION: 03/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:43

SUBJECT'S LAST NAME: MCCARTT

DATE/TIME OF TEST: 03/24/2017 22:13:57

Diagnostics Check	OK	22:05	
Air Blank	0.000	22:05	
Control Test	0.082	22:06	
Air Blank	0.000	22:06	
Subject Sample	# 1	22:09	No Sample Provided
Breath Volume	0.000	22:09	
Air Blank	0.000	22:10	
Air Blank	0.000	22:12	
Subject Sample	# 2	22:12	Subject Test Refused
Breath Volume	0.000	22:12	
Air Blank	0.000	22:13	
Control Test	0.081	22:13	
Air Blank	0.000	22:13	
Diagnostics Check	OK	22:13	

Data Download Date/Time: 04/21/2017 08:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSON, MATTHEW, P
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: MILLS, D
ARRESTING OFFICER AGENCY: Orange Park PD
LAST AGENCY INSPECTION: 03/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: PEETS

DATE/TIME OF TEST: 03/26/2017 03:39:37

Diagnostics Check	OK	03:31	
Air Blank	0.000	03:31	
Control Test	0.081	03:31	
Air Blank	0.000	03:32	
Subject Sample	# 1	03:35	No Sample Provided
Breath Volume	0.000	03:35	
Air Blank	0.000	03:36	
Air Blank	0.000	03:38	
Subject Sample	# 2	03:38	Subject Test Refused
Breath Volume	0.000	03:38	
Air Blank	0.000	03:38	
Control Test	0.081	03:39	
Air Blank	0.000	03:39	
Diagnostics Check	OK	03:39	

Data Download Date/Time: 04/21/2017 08:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANTIAGO,DEPUTY,
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: SOLIS,DEPUTY
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 03/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1276080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 05:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:10

SUBJECT'S LAST NAME: B625421767450

DATE/TIME OF TEST: 03/30/2017 06:46:23

Diagnostics Check	OK	06:44	
Air Blank	0.000	06:44	
Control Test	0.080	06:44	
Air Blank	0.000	06:45	
Subject Sample	# 1	06:45	Sequence Aborted
Breath Volume	0.000	06:45	
Air Blank	0.000	06:46	

Data Download Date/Time: 04/21/2017 08:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001471

INSTRUMENT REGISTERED TO: PATRICK AFB SFS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OUVerson,JESSE,T
BREATH TEST OPERATOR AGENCY: Patrick AFB PD
ARRESTING OFFICER: BEASLEY,SHAUN
ARRESTING OFFICER AGENCY: Patrick AFB PD
LAST AGENCY INSPECTION: 02/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1841081
DRY GAS EXPIRY DATE: 04/08/2018
VIOLATION CODE: ZERO TOL
ARREST TIME: 11:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:12

SUBJECT'S LAST NAME: ROBERTSON

DATE/TIME OF TEST: 03/03/2017 11:39:10

Diagnostics Check	OK	11:37	
Air Blank	0.000	11:38	
Control Test	0.032	11:38	Control Outside Tolerance
Air Blank	0.000	11:39	

Data Download Date/Time: 03/08/2017 16:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001471

INSTRUMENT REGISTERED TO: PATRICK AFB SFS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OUVerson,JESSE,T
BREATH TEST OPERATOR AGENCY: Patrick AFB PD
ARRESTING OFFICER: BEASLEY,SHAUN
ARRESTING OFFICER AGENCY: Patrick AFB PD
LAST AGENCY INSPECTION: 02/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1841081
DRY GAS EXPIRY DATE: 04/08/2018
VIOLATION CODE: OTHER
ARREST TIME: 11:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:12

SUBJECT'S LAST NAME: ROBERTSON

DATE/TIME OF TEST: 03/03/2017 11:42:32

Diagnostics Check	OK	11:41	
Air Blank	0.000	11:41	
Control Test	0.061	11:41	Control Outside Tolerance
Air Blank	0.000	11:42	

Data Download Date/Time: 03/08/2017 16:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001471

INSTRUMENT REGISTERED TO: PATRICK AFB SFS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OUVerson,JESSE,T
BREATH TEST OPERATOR AGENCY: Patrick AFB PD
ARRESTING OFFICER: BEASLEY,SHAUN
ARRESTING OFFICER AGENCY: Patrick AFB PD
LAST AGENCY INSPECTION: 02/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1841081
DRY GAS EXPIRY DATE: 04/08/2018
VIOLATION CODE: OTHER
ARREST TIME: 11:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:12

SUBJECT'S LAST NAME: ROBERTSON

DATE/TIME OF TEST: 03/03/2017 11:45:45

Diagnostics Check	OK	11:44	
Air Blank	0.000	11:44	
Control Test	0.044	11:45	Control Outside Tolerance
Air Blank	0.000	11:45	

Data Download Date/Time: 03/08/2017 16:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001471

INSTRUMENT REGISTERED TO: PATRICK AFB SFS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUNDEEN,SAMANTHA,L
BREATH TEST OPERATOR AGENCY: Patrick AFB PD
ARRESTING OFFICER: KLINE,MATTHEW
ARRESTING OFFICER AGENCY: Patrick AFB PD
LAST AGENCY INSPECTION: 03/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1841081
DRY GAS EXPIRY DATE: 04/08/2018
VIOLATION CODE: DUI
ARREST TIME: 20:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:59

SUBJECT'S LAST NAME: LUNDEEN

DATE/TIME OF TEST: 03/18/2017 21:49:23

Diagnostics Check	OK	21:41
Air Blank	0.000	21:42
Control Test	0.083	21:42
Air Blank	0.000	21:43
Subject Sample # 1	0.278	21:44
Breath Volume	2.503	21:44
Air Blank	0.000	21:45
Air Blank	0.000	21:47
Subject Sample # 2	0.263	21:47
Breath Volume	1.488	21:47
Air Blank	0.000	21:48
Control Test	0.077	21:48
Air Blank	0.000	21:49
Diagnostics Check	OK	21:49

Data Download Date/Time: 04/07/2017 14:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001649

INSTRUMENT REGISTERED TO: KENNETH CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VIENO,LT.MICHAEL,J
BREATH TEST OPERATOR AGENCY: Kenneth City PD
ARRESTING OFFICER: BRUNER,SEAN
ARRESTING OFFICER AGENCY: Kenneth City PD
LAST AGENCY INSPECTION: 02/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A3
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: BORSELLINO

DATE/TIME OF TEST: 03/21/2017 22:31:09

Diagnostics Check	OK	22:24
Air Blank	0.000	22:25
Control Test	0.079	22:25
Air Blank	0.000	22:25
Subject Sample # 1	0.107	22:26
Breath Volume	2.187	22:26
Air Blank	0.000	22:27
Air Blank	0.000	22:28
Subject Sample # 2	0.106	22:29
Breath Volume	2.542	22:29
Air Blank	0.000	22:30
Control Test	0.076	22:30
Air Blank	0.000	22:31
Diagnostics Check	OK	22:31

Data Download Date/Time: 03/22/2017 22:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001654

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIEIGA,FRANK,
BREATH TEST OPERATOR AGENCY: North Bay Village PD
ARRESTING OFFICER: ANGULO,FRANK
ARRESTING OFFICER AGENCY: North Bay Village PD
LAST AGENCY INSPECTION: 02/23/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 19:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:11

SUBJECT'S LAST NAME: MEDINA SAMON

DATE/TIME OF TEST: 03/04/2017 20:42:06

Diagnostics Check	OK	20:31
Air Blank	0.000	20:31
Control Test	0.081	20:32
Air Blank	0.000	20:32
Subject Sample # 1	0.145	20:33
Breath Volume	1.468	20:33
Air Blank	0.000	20:34
Air Blank	0.000	20:36
Subject Sample # 2	0.172	20:36
Breath Volume	1.449	20:36
Air Blank	0.000	20:37
Air Blank	0.000	20:39
Subject Sample # 3	0.169	20:40
Breath Volume	2.062	20:40
Air Blank	0.000	20:41
Control Test	0.080	20:41
Air Blank	0.000	20:42
Diagnostics Check	OK	20:42

Data Download Date/Time: 03/29/2017 17:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001656

INSTRUMENT REGISTERED TO: NORTH MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PINEDA,MIGUEL,A
BREATH TEST OPERATOR AGENCY: North Miami PD
ARRESTING OFFICER: PINEDA,MIGUEL
ARRESTING OFFICER AGENCY: North Miami PD
LAST AGENCY INSPECTION: 02/14/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06415080A2
DRY GAS EXPIRY DATE: 04/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:48

SUBJECT'S LAST NAME: JACKSON THOMAS

DATE/TIME OF TEST: 03/05/2017 03:28:45

Diagnostics Check	OK	03:25	
Air Blank	0.000	03:26	
Control Test	0.078	03:26	
Air Blank	0.000	03:27	
Subject Sample	# 1	03:27	Subject Test Refused
Breath Volume	0.000	03:27	
Air Blank	0.000	03:27	
Control Test	0.079	03:28	
Air Blank	0.000	03:28	
Diagnostics Check	OK	03:28	

Data Download Date/Time: 03/09/2017 13:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LYONS,D,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: LYONS,D
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 02/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 05:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:16

SUBJECT'S LAST NAME: STANDLEY

DATE/TIME OF TEST: 03/05/2017 05:43:49

Diagnostics Check	OK	05:36
Air Blank	0.000	05:37
Control Test	0.080	05:37
Air Blank	0.000	05:38
Subject Sample # 1	0.152	05:38
Breath Volume	1.410	05:38
Air Blank	0.000	05:39
Air Blank	0.000	05:41
Subject Sample # 2	0.166	05:42
Breath Volume	2.343	05:42
Air Blank	0.000	05:42
Control Test	0.079	05:43
Air Blank	0.000	05:43
Diagnostics Check	OK	05:43

Data Download Date/Time: 03/27/2017 17:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LYONS,D,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: BEEKMAN,JA.
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 02/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 23:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:26

SUBJECT'S LAST NAME: BLANCAS

DATE/TIME OF TEST: 03/10/2017 23:51:21

Diagnostics Check	OK	23:47	
Air Blank	0.000	23:47	
Control Test	0.079	23:47	
Air Blank	0.000	23:48	
Subject Sample	# 1 0.188	23:49	
Breath Volume	4.988	23:49	
Air Blank		23:49	RFI Detect
Air Blank	0.000	23:49	
Control Test	0.079	23:50	
Air Blank	0.000	23:50	
Diagnostics Check	Fail	23:50	Voltage/Current Test
Air Blank	0.000	23:51	

Data Download Date/Time: 03/27/2017 17:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LYONS,D,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: BEEKMAN,JA.
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 02/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 23:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:26

SUBJECT'S LAST NAME: BLANCAS

DATE/TIME OF TEST: 03/11/2017 00:00:46

Diagnostics Check	OK	23:54
Air Blank	0.000	23:54
Control Test	0.079	23:55
Air Blank	0.000	23:55
Subject Sample # 1	0.171	23:56
Breath Volume	3.296	23:56
Air Blank	0.000	23:57
Air Blank	0.000	23:58
Subject Sample # 2	0.172	23:59
Breath Volume	3.238	23:59
Air Blank	0.000	23:59
Control Test	0.078	00:00
Air Blank	0.000	00:00
Diagnostics Check	OK	00:00

Data Download Date/Time: 03/27/2017 17:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LYONS,J,D,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: LYONS,D
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 02/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 23:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:19

SUBJECT'S LAST NAME: QUEZADA-LENDERBORG

DATE/TIME OF TEST: 03/11/2017 23:47:05

Diagnostics Check	OK	23:40
Air Blank	0.000	23:40
Control Test	0.079	23:40
Air Blank	0.000	23:41
Subject Sample # 1	0.181	23:42
Breath Volume	1.398	23:42
Air Blank	0.000	23:42
Air Blank	0.000	23:44
Subject Sample # 2	0.177	23:45
Breath Volume	1.699	23:45
Air Blank	0.000	23:46
Control Test	0.079	23:46
Air Blank	0.000	23:47
Diagnostics Check	OK	23:47

Data Download Date/Time: 03/27/2017 17:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LYONS,D,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: LYONS,D
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 02/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 05:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:35

SUBJECT'S LAST NAME: BEDEL

DATE/TIME OF TEST: 03/12/2017 06:01:54

Diagnostics Check	OK	05:55
Air Blank	0.000	05:55
Control Test	0.079	05:56
Air Blank	0.000	05:56
Subject Sample # 1	0.166	05:57
Breath Volume	1.847	05:57
Air Blank	0.000	05:58
Air Blank	0.000	05:59
Subject Sample # 2	0.172	06:00
Breath Volume	1.550	06:00
Air Blank	0.000	06:00
Control Test	0.078	06:01
Air Blank	0.000	06:01
Diagnostics Check	OK	06:01

Data Download Date/Time: 03/27/2017 17:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LYONS,D,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: LYONS,D
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 02/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 00:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:37

SUBJECT'S LAST NAME: LEGERTON

DATE/TIME OF TEST: 03/15/2017 01:04:36

Diagnostics Check	OK	00:58	
Air Blank	0.000	00:59	
Control Test	0.080	00:59	
Air Blank	0.000	01:00	
Subject Sample	# 1 0.096	01:03	Volume Not Met
Breath Volume	0.503	01:03	
Air Blank		01:03	Purge Fail
Air Blank	0.000	01:04	

Data Download Date/Time: 03/27/2017 17:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LYONS,D,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: LYONS,D
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 02/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 00:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:37

SUBJECT'S LAST NAME: LEGERTON

DATE/TIME OF TEST: 03/15/2017 01:14:45

Diagnostics Check	OK	01:05
Air Blank	0.000	01:06
Control Test	0.077	01:06
Air Blank	0.000	01:07
Subject Sample # 1	0.112	01:08
Breath Volume	1.234	01:08
Air Blank	0.000	01:09
Air Blank	0.000	01:10
Subject Sample # 2	0.106	01:13
Breath Volume	1.484	01:13
Air Blank	0.000	01:13
Control Test	0.078	01:14
Air Blank	0.000	01:14
Diagnostics Check	OK	01:14

Data Download Date/Time: 03/27/2017 17:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LYONS,D,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: LYONS,D
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 02/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 22:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:15

SUBJECT'S LAST NAME: TRACEY

DATE/TIME OF TEST: 03/17/2017 22:47:40

Diagnostics Check	OK	22:35	
Air Blank	0.000	22:35	
Control Test	0.080	22:36	
Air Blank	0.000	22:36	
Subject Sample	# 1 0.131	22:39	Volume Not Met
Breath Volume	0.824	22:39	
Air Blank	0.000	22:40	
Air Blank	0.000	22:42	
Subject Sample	# 2 0.135	22:42	
Breath Volume	1.175	22:42	
Air Blank	0.000	22:43	
Air Blank	0.000	22:45	
Subject Sample	# 3 0.124	22:45	
Breath Volume	1.753	22:45	
Air Blank	0.000	22:46	
Control Test	0.079	22:47	
Air Blank	0.000	22:47	
Diagnostics Check	OK	22:47	

Data Download Date/Time: 03/27/2017 17:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LYONS,D,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: CAMPOS,M
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 02/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 01:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:12

SUBJECT'S LAST NAME: LARES BOHORQUEZ

DATE/TIME OF TEST: 03/19/2017 01:39:45

Diagnostics Check	OK	01:32
Air Blank	0.000	01:33
Control Test	0.080	01:33
Air Blank	0.000	01:33
Subject Sample # 1	0.151	01:34
Breath Volume	2.101	01:34
Air Blank	0.000	01:35
Air Blank	0.000	01:37
Subject Sample # 2	0.150	01:38
Breath Volume	1.933	01:38
Air Blank	0.000	01:38
Control Test	0.078	01:39
Air Blank	0.000	01:39
Diagnostics Check	OK	01:39

Data Download Date/Time: 03/27/2017 17:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LYONS,D,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: LYONS,D
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 02/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 03:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:29

SUBJECT'S LAST NAME: JAMES

DATE/TIME OF TEST: 03/19/2017 03:59:23

Diagnostics Check	OK	03:51
Air Blank	0.000	03:51
Control Test	0.080	03:52
Air Blank	0.000	03:52
Subject Sample # 1	0.000	03:55
Breath Volume	1.132	03:55
Air Blank	0.000	03:55
Air Blank	0.000	03:57
Subject Sample # 2	0.000	03:58
Breath Volume	2.566	03:58
Air Blank	0.000	03:58
Control Test	0.078	03:58
Air Blank	0.000	03:59
Diagnostics Check	OK	03:59

Data Download Date/Time: 03/27/2017 17:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LYONS,D,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: LYONS,D
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 02/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:28

SUBJECT'S LAST NAME: LUENEBURG

DATE/TIME OF TEST: 03/24/2017 00:09:14

Diagnostics Check	OK	23:59
Air Blank	0.000	23:59
Control Test	0.079	00:00
Air Blank	0.000	00:00
Subject Sample # 1	0.167	00:03
Breath Volume	1.433	00:03
Air Blank	0.000	00:04
Air Blank	0.000	00:05
Subject Sample # 2	0.166	00:07
Breath Volume	1.921	00:07
Air Blank	0.000	00:08
Control Test	0.077	00:08
Air Blank	0.000	00:09
Diagnostics Check	OK	00:09

Data Download Date/Time: 03/27/2017 17:31