

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000868

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN-REYES,JUAN,C
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: ZAKARIADZE,GEORGE
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 03/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 562838
DRY GAS EXPIRY DATE: 09/06/2019
VIOLATION CODE: DUI
ARREST TIME: 21:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: QUINTON

DATE/TIME OF TEST: 03/09/2017 22:17:41

Diagnostics Check	OK	22:10
Air Blank	0.000	22:11
Control Test	0.081	22:11
Air Blank	0.000	22:12
Subject Sample # 1	0.167	22:13
Breath Volume	2.882	22:13
Air Blank	0.000	22:13
Air Blank	0.000	22:15
Subject Sample # 2	0.162	22:16
Breath Volume	2.781	22:16
Air Blank	0.000	22:16
Control Test	0.080	22:17
Air Blank	0.000	22:17
Diagnostics Check	OK	22:17

Data Download Date/Time: 04/29/2017 09:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000868

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOPP,SETH,M
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: HOPP,SETH
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 03/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 562838
DRY GAS EXPIRY DATE: 09/06/2019
VIOLATION CODE: DUI
ARREST TIME: 23:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:53

SUBJECT'S LAST NAME: QUINN

DATE/TIME OF TEST: 03/11/2017 01:02:21

Diagnostics Check	OK	00:59	
Air Blank	0.000	00:59	
Control Test	0.080	01:00	
Air Blank	0.000	01:00	
Subject Sample	# 1	01:00	Subject Test Refused
Breath Volume	0.000	01:00	
Air Blank	0.000	01:01	
Control Test	0.081	01:01	
Air Blank	0.000	01:02	
Diagnostics Check	OK	01:02	

Data Download Date/Time: 04/29/2017 09:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000868

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLYTH,KYLE,A
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: BLYTH,KYLE
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 03/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 562838
DRY GAS EXPIRY DATE: 09/06/2019
VIOLATION CODE: DUI
ARREST TIME: 22:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:46

SUBJECT'S LAST NAME: HOWARD

DATE/TIME OF TEST: 03/22/2017 23:09:28

Diagnostics Check	OK	23:06	
Air Blank	0.000	23:07	
Control Test	0.080	23:07	
Air Blank	0.000	23:07	
Subject Sample	# 1	23:08	Subject Test Refused
Breath Volume	0.000	23:08	
Air Blank	0.000	23:08	
Control Test	0.081	23:08	
Air Blank	0.000	23:09	
Diagnostics Check	OK	23:09	

Data Download Date/Time: 04/29/2017 09:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000868

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN-REYES, JUAN, C
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: DANIELS, WILLIAM
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 03/04/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 562838
DRY GAS EXPIRY DATE: 09/06/2019
VIOLATION CODE: DUI
ARREST TIME: 01:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: CZETTLER

DATE/TIME OF TEST: 03/28/2017 02:14:37

Diagnostics Check	OK	02:10	
Air Blank	0.000	02:11	
Control Test	0.081	02:11	
Air Blank	0.000	02:12	
Subject Sample	# 1	02:13	Subject Test Refused
Breath Volume	0.000	02:13	
Air Blank	0.000	02:13	
Control Test	0.081	02:14	
Air Blank	0.000	02:14	
Diagnostics Check	OK	02:14	

Data Download Date/Time: 04/29/2017 09:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000868

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORBETA,JORGE,
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: JORGE,ORBETA
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 03/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 562838
DRY GAS EXPIRY DATE: 09/06/2019
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:53

SUBJECT'S LAST NAME: INGRAM

DATE/TIME OF TEST: 03/31/2017 03:25:35

Diagnostics Check	OK	03:13	
Air Blank	0.000	03:13	
Control Test	0.082	03:14	
Air Blank	0.000	03:14	
Subject Sample	# 1	03:17	No Sample Provided
Breath Volume	0.000	03:17	
Air Blank	0.000	03:18	
Air Blank	0.000	03:20	
Subject Sample	# 2	0.102	03:20
Breath Volume	4.816	03:20	
Air Blank	0.000	03:21	
Air Blank	0.000	03:23	
Subject Sample	# 3	0.107	03:24
Breath Volume	4.433	03:24	
Air Blank	0.000	03:24	
Control Test	0.082	03:24	
Air Blank	0.000	03:25	
Diagnostics Check	OK	03:25	

Data Download Date/Time: 04/29/2017 09:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOPP,AUSTIN,S
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: MOELLER,JENNA
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 02/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14968
DRY GAS EXPIRY DATE: 08/06/2018
VIOLATION CODE: DUI
ARREST TIME: 22:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: PERIARD

DATE/TIME OF TEST: 03/01/2017 23:16:12

Diagnostics Check	OK	23:13	
Air Blank	0.000	23:13	
Control Test	0.081	23:14	
Air Blank	0.000	23:14	
Subject Sample	# 1	23:14	Subject Test Refused
Breath Volume	0.000	23:14	
Air Blank	0.000	23:15	
Control Test	0.080	23:15	
Air Blank	0.000	23:16	
Diagnostics Check	OK	23:16	

Data Download Date/Time: 03/02/2017 09:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,ALEX,
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: RODRIGUEZ,ALEX
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 03/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14968
DRY GAS EXPIRY DATE: 08/06/2018
VIOLATION CODE: DUI
ARREST TIME: 18:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:18

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 03/03/2017 19:53:58

Diagnostics Check		OK	19:40	
Air Blank		0.000	19:40	
Control Test		0.081	19:41	
Air Blank		0.000	19:41	
Subject Sample	# 1	0.231	19:44	Volume Not Met
Breath Volume		0.289	19:44	
Air Blank		0.000	19:45	
Air Blank		0.000	19:47	
Subject Sample	# 2	0.253	19:49	
Breath Volume		1.539	19:49	
Air Blank		0.000	19:50	
Air Blank		0.000	19:51	
Subject Sample	# 3		19:52	Subject Test Refused
Breath Volume		0.000	19:52	
Air Blank		0.000	19:53	
Control Test		0.079	19:53	
Air Blank		0.000	19:53	
Diagnostics Check		OK	19:53	

Data Download Date/Time: 04/27/2017 13:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETZ,FRANK,L
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: KEOHANE,NIKO
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 03/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14968
DRY GAS EXPIRY DATE: 08/06/2018
VIOLATION CODE: DUI
ARREST TIME: 13:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:08

SUBJECT'S LAST NAME: MIZZONI

DATE/TIME OF TEST: 03/04/2017 14:40:04

Diagnostics Check	OK	14:28
Air Blank	0.000	14:28
Control Test	0.082	14:29
Air Blank	0.000	14:29
Subject Sample # 1	0.176	14:32
Breath Volume	2.371	14:32
Air Blank	0.000	14:33
Air Blank	0.000	14:34
Subject Sample # 2	0.153	14:35
Breath Volume	1.390	14:35
Air Blank	0.000	14:36
Air Blank	0.000	14:38
Subject Sample # 3	0.177	14:38
Breath Volume	1.222	14:38
Air Blank	0.000	14:39
Control Test	0.080	14:39
Air Blank	0.000	14:39
Diagnostics Check	OK	14:40

Data Download Date/Time: 04/27/2017 13:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CURRUL,ROBERT,G
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: YOUNG,JESSE
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 03/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14968
DRY GAS EXPIRY DATE: 08/06/2018
VIOLATION CODE: DUI
ARREST TIME: 18:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:00

SUBJECT'S LAST NAME: WEEKS

DATE/TIME OF TEST: 03/06/2017 19:31:53

Diagnostics Check	OK	19:24
Air Blank	0.000	19:25
Control Test	0.082	19:25
Air Blank	0.000	19:25
Subject Sample # 1	0.000	19:27
Breath Volume	1.890	19:27
Air Blank	0.000	19:27
Air Blank	0.000	19:29
Subject Sample # 2	0.000	19:30
Breath Volume	1.101	19:30
Air Blank	0.000	19:30
Control Test	0.081	19:31
Air Blank	0.000	19:31
Diagnostics Check	OK	19:31

Data Download Date/Time: 04/27/2017 13:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCGEE,JOHN,W
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: MCGEE,JW
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 03/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14968
DRY GAS EXPIRY DATE: 08/06/2018
VIOLATION CODE: DUI
ARREST TIME: 14:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:23

SUBJECT'S LAST NAME: KLEPASKIS

DATE/TIME OF TEST: 03/08/2017 14:50:57

Diagnostics Check	OK	14:43
Air Blank	0.000	14:44
Control Test	0.081	14:44
Air Blank	0.000	14:44
Subject Sample # 1	0.320	14:45
Breath Volume	1.355	14:45
Air Blank	0.000	14:46
Air Blank	0.000	14:47
Subject Sample # 2	0.326	14:49
Breath Volume	1.570	14:49
Air Blank	0.000	14:50
Control Test	0.079	14:50
Air Blank	0.000	14:50
Diagnostics Check	OK	14:50

Data Download Date/Time: 04/27/2017 13:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,JESSE,A
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: DEHANAS,BRETT
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 03/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14968
DRY GAS EXPIRY DATE: 08/06/2018
VIOLATION CODE: DUI
ARREST TIME: 00:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:11

SUBJECT'S LAST NAME: SAAD

DATE/TIME OF TEST: 03/10/2017 02:03:38

Diagnostics Check	OK	02:00	
Air Blank	0.000	02:00	
Control Test	0.081	02:01	
Air Blank	0.000	02:01	
Subject Sample	# 1	02:02	Subject Test Refused
Breath Volume	0.000	02:02	
Air Blank	0.000	02:02	
Control Test	0.081	02:03	
Air Blank	0.000	02:03	
Diagnostics Check	OK	02:03	

Data Download Date/Time: 04/27/2017 13:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,JESSE,A
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: YOUNG,JESSE
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 03/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14968
DRY GAS EXPIRY DATE: 08/06/2018
VIOLATION CODE: DUI
ARREST TIME: 01:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:06

SUBJECT'S LAST NAME: DUCLOS

DATE/TIME OF TEST: 03/10/2017 02:41:07

Diagnostics Check	OK	02:29	
Air Blank	0.000	02:30	
Control Test	0.079	02:30	
Air Blank	0.000	02:31	
Subject Sample	# 1	02:34	No Sample Provided
Breath Volume	0.000	02:34	
Air Blank	0.000	02:34	
Air Blank	0.000	02:36	
Subject Sample	# 2	02:39	No Sample Provided
Breath Volume	0.000	02:39	
Air Blank	0.000	02:40	
Control Test	0.082	02:40	
Air Blank	0.000	02:41	
Diagnostics Check	OK	02:41	

Data Download Date/Time: 04/27/2017 13:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,REINIER,
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: LOPEZ,ANTHONY
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 03/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14968
DRY GAS EXPIRY DATE: 08/06/2018
VIOLATION CODE: DUI
ARREST TIME: 04:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: MCANALLY

DATE/TIME OF TEST: 03/11/2017 05:12:29

Diagnostics Check	OK	05:02	
Air Blank	0.000	05:02	
Control Test	0.081	05:03	
Air Blank	0.000	05:03	
Subject Sample	# 1 0.294	05:05	
Breath Volume	1.808	05:05	
Air Blank	0.000	05:06	
Air Blank	0.000	05:07	
Subject Sample	# 2	05:10	Subject Test Refused
Breath Volume	0.855	05:10	
Air Blank	0.000	05:11	
Control Test	0.081	05:11	
Air Blank	0.000	05:12	
Diagnostics Check	OK	05:12	

Data Download Date/Time: 04/27/2017 13:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,JESSE,A
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: YOUNG,JESSE
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 03/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14968
DRY GAS EXPIRY DATE: 08/06/2018
VIOLATION CODE: DUI
ARREST TIME: 20:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:07

SUBJECT'S LAST NAME: MARCINEK

DATE/TIME OF TEST: 03/11/2017 21:36:01

Diagnostics Check	OK	21:33	
Air Blank	0.000	21:33	
Control Test	0.081	21:33	
Air Blank	0.000	21:34	
Subject Sample	# 1	21:34	Subject Test Refused
Breath Volume	0.000	21:34	
Air Blank	0.000	21:35	
Control Test	0.080	21:35	
Air Blank	0.000	21:35	
Diagnostics Check	OK	21:36	

Data Download Date/Time: 04/27/2017 13:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,JESSE,A
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: YOUNG,JESSE
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 03/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14968
DRY GAS EXPIRY DATE: 08/06/2018
VIOLATION CODE: DUI
ARREST TIME: 03:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: SWAYER

DATE/TIME OF TEST: 03/12/2017 04:08:56

Diagnostics Check	OK	04:05	
Air Blank	0.000	04:06	
Control Test	0.082	04:06	
Air Blank	0.000	04:07	
Subject Sample	# 1	04:07	Subject Test Refused
Breath Volume	0.000	04:07	
Air Blank	0.000	04:08	
Control Test	0.082	04:08	
Air Blank	0.000	04:08	
Diagnostics Check	OK	04:08	

Data Download Date/Time: 04/27/2017 13:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,ALEX,
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: RODRIGUEZ,ALEX
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 03/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14968
DRY GAS EXPIRY DATE: 08/06/2018
VIOLATION CODE: DUI
ARREST TIME: 03:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:38

SUBJECT'S LAST NAME: LAGOS

DATE/TIME OF TEST: 03/15/2017 04:06:18

Diagnostics Check	OK	04:00
Air Blank	0.000	04:00
Control Test	0.082	04:00
Air Blank	0.000	04:01
Subject Sample # 1	0.184	04:01
Breath Volume	2.964	04:01
Air Blank	0.000	04:02
Air Blank	0.000	04:04
Subject Sample # 2	0.182	04:04
Breath Volume	2.859	04:04
Air Blank	0.000	04:05
Control Test	0.081	04:05
Air Blank	0.000	04:06
Diagnostics Check	OK	04:06

Data Download Date/Time: 04/27/2017 13:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,JESSE,A
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: YOUNG,JESSE
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 03/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14968
DRY GAS EXPIRY DATE: 08/06/2018
VIOLATION CODE: DUI
ARREST TIME: 19:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:10

SUBJECT'S LAST NAME: BURGESS

DATE/TIME OF TEST: 03/15/2017 20:47:14

Diagnostics Check	OK	20:35	
Air Blank	0.000	20:36	
Control Test	0.082	20:36	
Air Blank	0.000	20:37	
Subject Sample	# 1	20:40	No Sample Provided
Breath Volume	0.000	20:40	
Air Blank	0.000	20:40	
Air Blank	0.000	20:42	
Subject Sample	# 2	20:45	No Sample Provided
Breath Volume	0.000	20:45	
Air Blank	0.000	20:46	
Control Test	0.082	20:46	
Air Blank	0.000	20:47	
Diagnostics Check	OK	20:47	

Data Download Date/Time: 04/27/2017 13:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,ALEX,
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: RODRIGUEZ,ALEX
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 03/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14968
DRY GAS EXPIRY DATE: 08/06/2018
VIOLATION CODE: DUI
ARREST TIME: 18:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:01

SUBJECT'S LAST NAME: BUNN

DATE/TIME OF TEST: 03/19/2017 19:28:26

Diagnostics Check	OK	19:23	
Air Blank	0.000	19:24	
Control Test	0.082	19:24	
Air Blank	0.000	19:24	
Subject Sample	# 1	19:27	Subject Test Refused
Breath Volume	0.000	19:27	
Air Blank	0.000	19:27	
Control Test	0.082	19:27	
Air Blank	0.000	19:28	
Diagnostics Check	OK	19:28	

Data Download Date/Time: 04/27/2017 13:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CALHOUN, EVAN, R
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: CALHOUN, EVAN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 03/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14968
DRY GAS EXPIRY DATE: 08/06/2018
VIOLATION CODE: DUI
ARREST TIME: 20:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:17

SUBJECT'S LAST NAME: BELL

DATE/TIME OF TEST: 03/22/2017 21:45:39

Diagnostics Check	OK	21:38
Air Blank	0.000	21:38
Control Test	0.082	21:38
Air Blank	0.000	21:39
Subject Sample # 1	0.139	21:40
Breath Volume	1.449	21:40
Air Blank	0.000	21:41
Air Blank	0.000	21:42
Subject Sample # 2	0.137	21:44
Breath Volume	1.390	21:44
Air Blank	0.000	21:44
Control Test	0.081	21:45
Air Blank	0.000	21:45
Diagnostics Check	OK	21:45

Data Download Date/Time: 04/27/2017 13:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KEOHANE,KUNIKO,
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: KEOHANE,KUNIKO
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 03/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14968
DRY GAS EXPIRY DATE: 08/06/2018
VIOLATION CODE: DUI
ARREST TIME: 15:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:15

SUBJECT'S LAST NAME: LILEKS

DATE/TIME OF TEST: 03/27/2017 17:05:45

Diagnostics Check	OK	17:02	
Air Blank	0.000	17:03	
Control Test	0.082	17:03	
Air Blank	0.000	17:04	
Subject Sample	# 1	17:04	Subject Test Refused
Breath Volume	0.000	17:04	
Air Blank	0.000	17:04	
Control Test	0.081	17:05	
Air Blank	0.000	17:05	
Diagnostics Check	OK	17:05	

Data Download Date/Time: 04/27/2017 13:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOPP,AUSTIN,S
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: LOPEZ,ANTHONY
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 03/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14968
DRY GAS EXPIRY DATE: 08/06/2018
VIOLATION CODE: DUI
ARREST TIME: 01:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:32

SUBJECT'S LAST NAME: GLATZ

DATE/TIME OF TEST: 03/30/2017 02:59:48

Diagnostics Check	OK	02:53
Air Blank	0.000	02:53
Control Test	0.081	02:53
Air Blank	0.000	02:54
Subject Sample # 1	0.227	02:54
Breath Volume	2.722	02:54
Air Blank	0.000	02:55
Air Blank	0.000	02:57
Subject Sample # 2	0.226	02:58
Breath Volume	2.027	02:58
Air Blank	0.000	02:58
Control Test	0.080	02:59
Air Blank	0.000	02:59
Diagnostics Check	OK	02:59

Data Download Date/Time: 04/27/2017 13:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,JESUS,J
BREATH TEST OPERATOR AGENCY: Coral Gables PD
ARRESTING OFFICER: TAHA,KHALED
ARRESTING OFFICER AGENCY: Coral Gables PD
LAST AGENCY INSPECTION: 02/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:46

SUBJECT'S LAST NAME: GOMEZFUNDORA

DATE/TIME OF TEST: 03/01/2017 02:10:51

Diagnostics Check	OK	02:07	
Air Blank	0.000	02:07	
Control Test	0.082	02:08	
Air Blank	0.000	02:08	
Subject Sample	# 1	02:09	Subject Test Refused
Breath Volume	0.000	02:09	
Air Blank	0.000	02:09	
Control Test	0.081	02:10	
Air Blank	0.000	02:10	
Diagnostics Check	OK	02:10	

Data Download Date/Time: 03/29/2017 12:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAHA,KHALED,
BREATH TEST OPERATOR AGENCY: Coral Gables PD
ARRESTING OFFICER: GRIGG,DAVID
ARRESTING OFFICER AGENCY: Coral Gables PD
LAST AGENCY INSPECTION: 02/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:24

SUBJECT'S LAST NAME: DANG

DATE/TIME OF TEST: 03/06/2017 02:47:15

Diagnostics Check	OK	02:44	
Air Blank	0.000	02:44	
Control Test	0.081	02:45	
Air Blank	0.000	02:45	
Subject Sample	# 1	02:45	Subject Test Refused
Breath Volume	0.000	02:45	
Air Blank	0.000	02:46	
Control Test	0.081	02:46	
Air Blank	0.000	02:47	
Diagnostics Check	OK	02:47	

Data Download Date/Time: 03/29/2017 12:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,JESUS,J
BREATH TEST OPERATOR AGENCY: Coral Gables PD
ARRESTING OFFICER: TAHA,KHALED
ARRESTING OFFICER AGENCY: Coral Gables PD
LAST AGENCY INSPECTION: 02/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:53

SUBJECT'S LAST NAME: LLANOS

DATE/TIME OF TEST: 03/08/2017 02:19:07

Diagnostics Check	OK	02:14	
Air Blank	0.000	02:14	
Control Test	0.081	02:15	
Air Blank	0.000	02:15	
Subject Sample	# 1	02:17	Subject Test Refused
Breath Volume	0.000	02:17	
Air Blank	0.000	02:18	
Control Test	0.081	02:18	
Air Blank	0.000	02:19	
Diagnostics Check	OK	02:19	

Data Download Date/Time: 03/29/2017 12:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAHA,KHALED,
BREATH TEST OPERATOR AGENCY: Coral Gables PD
ARRESTING OFFICER: TAHA,KHALED
ARRESTING OFFICER AGENCY: Coral Gables PD
LAST AGENCY INSPECTION: 02/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: BRAKA

DATE/TIME OF TEST: 03/12/2017 03:27:52

Diagnostics Check	OK	03:21
Air Blank	0.000	03:22
Control Test	0.080	03:22
Air Blank	0.000	03:22
Subject Sample # 1	0.138	03:23
Breath Volume	1.910	03:23
Air Blank	0.000	03:24
Air Blank	0.000	03:25
Subject Sample # 2	0.143	03:26
Breath Volume	2.523	03:26
Air Blank	0.000	03:26
Control Test	0.080	03:27
Air Blank	0.000	03:27
Diagnostics Check	OK	03:27

Data Download Date/Time: 03/29/2017 12:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLORES,EMILIO,E
BREATH TEST OPERATOR AGENCY: Coral Gables PD
ARRESTING OFFICER: NAYA,CHRISTOPHER
ARRESTING OFFICER AGENCY: Coral Gables PD
LAST AGENCY INSPECTION: 02/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 07:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:00

SUBJECT'S LAST NAME: HIDALGO

DATE/TIME OF TEST: 03/18/2017 08:50:24

Diagnostics Check	OK	08:47	
Air Blank	0.000	08:47	
Control Test	0.081	08:48	
Air Blank	0.000	08:48	
Subject Sample	# 1	08:49	Subject Test Refused
Breath Volume	0.000	08:49	
Air Blank	0.000	08:49	
Control Test	0.081	08:49	
Air Blank	0.000	08:50	
Diagnostics Check	OK	08:50	

Data Download Date/Time: 03/29/2017 12:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000874

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VARGAS,D,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: VARGAS,D
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 02/23/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 18:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:40

SUBJECT'S LAST NAME: LAZO ROQUE

DATE/TIME OF TEST: 03/18/2017 19:15:50

Diagnostics Check	OK	19:09
Air Blank	0.000	19:10
Control Test	0.079	19:10
Air Blank	0.000	19:10
Subject Sample # 1	0.207	19:11
Breath Volume	2.976	19:11
Air Blank	0.000	19:12
Air Blank	0.000	19:13
Subject Sample # 2	0.194	19:14
Breath Volume	2.710	19:14
Air Blank	0.000	19:14
Control Test	0.079	19:15
Air Blank	0.000	19:15
Diagnostics Check	OK	19:15

Data Download Date/Time: 04/19/2017 12:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000874

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PASTOR,JOSE,F
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: DIAZ,LESLEY
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 02/23/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 21:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:25

SUBJECT'S LAST NAME: MAYORGA

DATE/TIME OF TEST: 03/18/2017 22:00:31

Diagnostics Check	OK	21:55	
Air Blank	0.000	21:55	
Control Test	0.080	21:56	
Air Blank	0.000	21:56	
Subject Sample	# 1	21:58	Subject Test Refused
Breath Volume	0.000	21:58	
Air Blank	0.000	21:59	
Control Test	0.080	21:59	
Air Blank	0.000	21:59	
Diagnostics Check	Fail	22:00	Temperature Regulation Test
Air Blank	0.000	22:00	

Data Download Date/Time: 04/19/2017 12:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000875

INSTRUMENT REGISTERED TO: CALHOUN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSON,JAMES,D
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: GUIFARRO,NORMA
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 03/09/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A3
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 05:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:45

SUBJECT'S LAST NAME: EDWARDS

DATE/TIME OF TEST: 03/11/2017 06:46:54

Diagnostics Check	OK	06:37
Air Blank	0.000	06:37
Control Test	0.079	06:37
Air Blank	0.000	06:38
Subject Sample # 1	0.181	06:39
Breath Volume	1.617	06:39
Air Blank	0.000	06:39
Air Blank	0.000	06:41
Subject Sample # 2	0.203	06:42
Breath Volume	2.988	06:42
Air Blank	0.000	06:42
Air Blank	0.000	06:44
Subject Sample # 3	0.204	06:45
Breath Volume	3.636	06:45
Air Blank	0.000	06:46
Control Test	0.079	06:46
Air Blank	0.000	06:46
Diagnostics Check	OK	06:46

Data Download Date/Time: 04/05/2017 10:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000878

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SINCORE,MICHAEL,A
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: SINCORE,MICHAEL
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 02/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1673300
DRY GAS EXPIRY DATE: 04/28/2017
VIOLATION CODE: DUI
ARREST TIME: 02:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: RAYMUNDOHERNANDEZ

DATE/TIME OF TEST: 03/05/2017 03:22:36

Diagnostics Check	OK	03:15
Air Blank	0.000	03:16
Control Test	0.082	03:16
Air Blank	0.000	03:17
Subject Sample # 1	0.266	03:17
Breath Volume	2.210	03:17
Air Blank	0.000	03:18
Air Blank	0.000	03:20
Subject Sample # 2	0.268	03:20
Breath Volume	2.964	03:20
Air Blank	0.000	03:21
Control Test	0.083	03:21
Air Blank	0.000	03:22
Diagnostics Check	OK	03:22

Data Download Date/Time: 03/29/2017 10:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000881

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,STEVEN,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: ADAMS,STEVEN
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 02/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 03/02/2017 02:06:09

Diagnostics Check	OK	02:00
Air Blank	0.000	02:00
Control Test	0.081	02:00
Air Blank	0.000	02:01
Subject Sample # 1	0.178	02:01
Breath Volume	1.230	02:01
Air Blank	0.000	02:02
Air Blank	0.000	02:04
Subject Sample # 2	0.179	02:04
Breath Volume	1.449	02:04
Air Blank	0.000	02:05
Control Test	0.079	02:05
Air Blank	0.000	02:06
Diagnostics Check	OK	02:06

Data Download Date/Time: 03/17/2017 12:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000881

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,JESSE,N
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: SMITH,JESSE
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 02/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 12:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:55

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 03/02/2017 13:29:28

Diagnostics Check	OK	13:22
Air Blank	0.000	13:23
Control Test	0.081	13:23
Air Blank	0.000	13:24
Subject Sample # 1	0.000	13:24
Breath Volume	1.917	13:24
Air Blank	0.000	13:25
Air Blank	0.000	13:27
Subject Sample # 2	0.000	13:28
Breath Volume	1.800	13:28
Air Blank	0.000	13:28
Control Test	0.080	13:28
Air Blank	0.000	13:29
Diagnostics Check	OK	13:29

Data Download Date/Time: 03/17/2017 12:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000881

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PACHECO,GEORGE,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: PACHECO,GEORGE
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 02/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:17

SUBJECT'S LAST NAME: LLAMBES

DATE/TIME OF TEST: 03/08/2017 01:52:47

Diagnostics Check	OK	01:46
Air Blank	0.000	01:46
Control Test	0.081	01:47
Air Blank	0.000	01:47
Subject Sample # 1	0.075	01:48
Breath Volume	3.132	01:48
Air Blank	0.000	01:49
Air Blank	0.000	01:50
Subject Sample # 2	0.077	01:51
Breath Volume	4.007	01:51
Air Blank	0.000	01:51
Control Test	0.080	01:52
Air Blank	0.000	01:52
Diagnostics Check	OK	01:52

Data Download Date/Time: 03/17/2017 12:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: CLOSIUS,RICHARD
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 03/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: DOMINGUEZ

DATE/TIME OF TEST: 03/02/2017 02:34:37

Diagnostics Check	OK	02:31	
Air Blank	0.000	02:31	
Control Test	0.078	02:32	
Air Blank	0.000	02:32	
Subject Sample	# 1	02:33	Subject Test Refused
Breath Volume	0.000	02:33	
Air Blank	0.000	02:33	
Control Test	0.079	02:33	
Air Blank	0.000	02:34	
Diagnostics Check	OK	02:34	

Data Download Date/Time: 04/06/2017 10:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 03/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:47

SUBJECT'S LAST NAME: WATSON

DATE/TIME OF TEST: 03/03/2017 05:22:01

Diagnostics Check	OK	05:11
Air Blank	0.000	05:12
Control Test	0.078	05:12
Air Blank	0.000	05:13
Subject Sample # 1	0.229	05:15
Breath Volume	1.882	05:15
Air Blank	0.000	05:16
Air Blank	0.000	05:18
Subject Sample # 2	0.222	05:20
Breath Volume	1.109	05:20
Air Blank	0.000	05:21
Control Test	0.078	05:21
Air Blank	0.000	05:21
Diagnostics Check	OK	05:22

Data Download Date/Time: 04/06/2017 10:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: FRANCO,GREGORY
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 03/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:21

SUBJECT'S LAST NAME: CASAYA

DATE/TIME OF TEST: 03/04/2017 03:48:16

Diagnostics Check	OK	03:42
Air Blank	0.000	03:42
Control Test	0.081	03:42
Air Blank	0.000	03:43
Subject Sample # 1	0.192	03:43
Breath Volume	3.171	03:43
Air Blank	0.000	03:44
Air Blank	0.000	03:46
Subject Sample # 2	0.194	03:46
Breath Volume	2.980	03:46
Air Blank	0.000	03:47
Control Test	0.080	03:47
Air Blank	0.000	03:48
Diagnostics Check	OK	03:48

Data Download Date/Time: 04/06/2017 10:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 03/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 03/04/2017 04:12:08

Diagnostics Check	OK	04:05
Air Blank	0.000	04:06
Control Test	0.080	04:06
Air Blank	0.000	04:06
Subject Sample # 1	0.176	04:07
Breath Volume	1.515	04:07
Air Blank	0.000	04:08
Air Blank	0.000	04:10
Subject Sample # 2	0.174	04:10
Breath Volume	2.007	04:10
Air Blank	0.000	04:11
Control Test	0.079	04:11
Air Blank	0.000	04:12
Diagnostics Check	OK	04:12

Data Download Date/Time: 04/06/2017 10:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: CLOSIUS,RICHARD
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 03/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:54

SUBJECT'S LAST NAME: CLOUSER

DATE/TIME OF TEST: 03/05/2017 02:18:19

Diagnostics Check	OK	02:15	
Air Blank	0.000	02:15	
Control Test	0.080	02:16	
Air Blank	0.000	02:16	
Subject Sample	# 1	02:16	Subject Test Refused
Breath Volume	0.000	02:16	
Air Blank	0.000	02:17	
Control Test	0.080	02:17	
Air Blank	0.000	02:18	
Diagnostics Check	OK	02:18	

Data Download Date/Time: 04/06/2017 10:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 03/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:06

SUBJECT'S LAST NAME: DUQUE

DATE/TIME OF TEST: 03/05/2017 04:32:14

Diagnostics Check	OK	04:29	
Air Blank	0.000	04:29	
Control Test	0.079	04:29	
Air Blank	0.000	04:30	
Subject Sample	# 1	04:30	Subject Test Refused
Breath Volume	0.000	04:30	
Air Blank	0.000	04:31	
Control Test	0.080	04:31	
Air Blank	0.000	04:32	
Diagnostics Check	OK	04:32	

Data Download Date/Time: 04/06/2017 10:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: CLOSIUS,RICHARD
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 03/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: CLAVEROL GIL

DATE/TIME OF TEST: 03/16/2017 00:13:02

Diagnostics Check	OK	00:10	
Air Blank	0.000	00:10	
Control Test	0.079	00:10	
Air Blank	0.000	00:11	
Subject Sample	# 1	00:11	Subject Test Refused
Breath Volume	0.000	00:11	
Air Blank	0.000	00:12	
Control Test	0.080	00:12	
Air Blank	0.000	00:12	
Diagnostics Check	OK	00:13	

Data Download Date/Time: 04/06/2017 10:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: CLOSIUS,RICHARD
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 03/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:25

SUBJECT'S LAST NAME: ESPINOSA SENRA

DATE/TIME OF TEST: 03/17/2017 04:58:44

Diagnostics Check	OK	04:51
Air Blank	0.000	04:52
Control Test	0.080	04:52
Air Blank	0.000	04:53
Subject Sample # 1	0.245	04:54
Breath Volume	3.554	04:54
Air Blank	0.000	04:55
Air Blank	0.000	04:56
Subject Sample # 2	0.247	04:57
Breath Volume	3.660	04:57
Air Blank	0.000	04:57
Control Test	0.079	04:58
Air Blank	0.000	04:58
Diagnostics Check	OK	04:58

Data Download Date/Time: 04/06/2017 10:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: CLOSIUS,RICHARD
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 03/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:06

SUBJECT'S LAST NAME: BEATON

DATE/TIME OF TEST: 03/18/2017 00:30:32

Diagnostics Check	OK	00:27	
Air Blank	0.000	00:28	
Control Test	0.080	00:28	
Air Blank	0.000	00:28	
Subject Sample	# 1	00:29	Subject Test Refused
Breath Volume	0.000	00:29	
Air Blank	0.000	00:29	
Control Test	0.080	00:29	
Air Blank	0.000	00:30	
Diagnostics Check	OK	00:30	

Data Download Date/Time: 04/06/2017 10:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: CLOSIUS,RICHARD
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 03/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:42

SUBJECT'S LAST NAME: ROBEY

DATE/TIME OF TEST: 03/23/2017 01:17:47

Diagnostics Check	OK	01:10
Air Blank	0.000	01:10
Control Test	0.079	01:11
Air Blank	0.000	01:11
Subject Sample # 1	0.215	01:13
Breath Volume	2.191	01:13
Air Blank	0.000	01:13
Air Blank	0.000	01:15
Subject Sample # 2	0.222	01:16
Breath Volume	2.710	01:16
Air Blank	0.000	01:16
Control Test	0.078	01:17
Air Blank	0.000	01:17
Diagnostics Check	OK	01:17

Data Download Date/Time: 04/06/2017 10:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: FRANCO,GREGORY
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 03/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: VALVERDE-MONTIEL

DATE/TIME OF TEST: 03/23/2017 04:26:07

Diagnostics Check	OK	04:23	
Air Blank	0.000	04:23	
Control Test	0.079	04:23	
Air Blank	0.000	04:24	
Subject Sample	# 1	04:24	Subject Test Refused
Breath Volume	0.000	04:24	
Air Blank	0.000	04:25	
Control Test	0.079	04:25	
Air Blank	0.000	04:26	
Diagnostics Check	OK	04:26	

Data Download Date/Time: 04/06/2017 10:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 03/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:13

SUBJECT'S LAST NAME: CALDERON

DATE/TIME OF TEST: 03/23/2017 05:42:37

Diagnostics Check	OK	05:36
Air Blank	0.000	05:37
Control Test	0.079	05:37
Air Blank	0.000	05:37
Subject Sample # 1	0.174	05:38
Breath Volume	1.710	05:38
Air Blank	0.000	05:38
Air Blank	0.000	05:40
Subject Sample # 2	0.171	05:41
Breath Volume	1.421	05:41
Air Blank	0.000	05:41
Control Test	0.079	05:41
Air Blank	0.000	05:42
Diagnostics Check	OK	05:42

Data Download Date/Time: 04/06/2017 10:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: CLOSIUS,RICHARD
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 03/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:37

SUBJECT'S LAST NAME: VELASQUEZ

DATE/TIME OF TEST: 03/25/2017 00:05:03

Diagnostics Check	OK	23:59
Air Blank	0.000	23:59
Control Test	0.079	23:59
Air Blank	0.000	00:00
Subject Sample # 1	0.088	00:00
Breath Volume	4.222	00:00
Air Blank	0.000	00:01
Air Blank	0.000	00:03
Subject Sample # 2	0.085	00:03
Breath Volume	4.242	00:03
Air Blank	0.000	00:04
Control Test	0.080	00:04
Air Blank	0.000	00:04
Diagnostics Check	OK	00:05

Data Download Date/Time: 04/06/2017 10:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRANCO,GREGORY,R
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: FRANCO,GREGORY
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 03/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:21

SUBJECT'S LAST NAME: PENA

DATE/TIME OF TEST: 03/25/2017 03:53:04

Diagnostics Check	OK	03:50	
Air Blank	0.000	03:50	
Control Test	0.079	03:51	
Air Blank	0.000	03:51	
Subject Sample	# 1	03:51	Subject Test Refused
Breath Volume	0.000	03:51	
Air Blank	0.000	03:52	
Control Test	0.079	03:52	
Air Blank	0.000	03:52	
Diagnostics Check	OK	03:53	

Data Download Date/Time: 04/06/2017 10:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRANCO,GREGORY,R
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: FRANCO,GREGORY
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 03/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: FIDALGO-CABRERA

DATE/TIME OF TEST: 03/26/2017 00:11:52

Diagnostics Check		OK	00:05
Air Blank		0.000	00:05
Control Test		0.079	00:06
Air Blank		0.000	00:06
Subject Sample	# 1	0.204	00:07
Breath Volume		2.101	00:07
Air Blank		0.000	00:07
Air Blank		0.000	00:09
Subject Sample	# 2	0.210	00:10
Breath Volume		2.742	00:10
Air Blank		0.000	00:10
Control Test		0.078	00:11
Air Blank		0.000	00:11
Diagnostics Check		OK	00:11

Data Download Date/Time: 04/06/2017 10:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 03/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:06

SUBJECT'S LAST NAME: PLETEAU

DATE/TIME OF TEST: 03/26/2017 02:35:54

Diagnostics Check	OK	02:29
Air Blank	0.000	02:30
Control Test	0.078	02:30
Air Blank	0.000	02:31
Subject Sample # 1	0.136	02:31
Breath Volume	1.625	02:31
Air Blank	0.000	02:32
Air Blank	0.000	02:34
Subject Sample # 2	0.137	02:34
Breath Volume	1.820	02:34
Air Blank	0.000	02:34
Control Test	0.079	02:35
Air Blank	0.000	02:35
Diagnostics Check	OK	02:35

Data Download Date/Time: 04/06/2017 10:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: CLOSIUS,RICHARD
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 03/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:03

SUBJECT'S LAST NAME: HALL

DATE/TIME OF TEST: 03/26/2017 05:28:27

Diagnostics Check	OK	05:25	
Air Blank	0.000	05:26	
Control Test	0.079	05:26	
Air Blank	0.000	05:26	
Subject Sample	# 1	05:27	Subject Test Refused
Breath Volume	0.000	05:27	
Air Blank	0.000	05:27	
Control Test	0.079	05:27	
Air Blank	0.000	05:28	
Diagnostics Check	OK	05:28	

Data Download Date/Time: 04/06/2017 10:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: TAHA,KHALID
ARRESTING OFFICER AGENCY: Coral Gables PD
LAST AGENCY INSPECTION: 03/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: MELTON

DATE/TIME OF TEST: 03/30/2017 02:38:45

Diagnostics Check	OK	02:32
Air Blank	0.000	02:32
Control Test	0.078	02:33
Air Blank	0.000	02:33
Subject Sample # 1	0.157	02:34
Breath Volume	1.828	02:34
Air Blank	0.000	02:35
Air Blank	0.000	02:36
Subject Sample # 2	0.153	02:37
Breath Volume	1.183	02:37
Air Blank	0.000	02:37
Control Test	0.078	02:38
Air Blank	0.000	02:38
Diagnostics Check	OK	02:38

Data Download Date/Time: 04/06/2017 10:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: CLOSIUS,RICHARD
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 03/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:06

SUBJECT'S LAST NAME: OLLIE

DATE/TIME OF TEST: 03/30/2017 04:32:24

Diagnostics Check	OK	04:26
Air Blank	0.000	04:26
Control Test	0.079	04:27
Air Blank	0.000	04:27
Subject Sample # 1	0.095	04:28
Breath Volume	1.480	04:28
Air Blank	0.000	04:28
Air Blank	0.000	04:30
Subject Sample # 2	0.091	04:30
Breath Volume	1.457	04:30
Air Blank	0.000	04:31
Control Test	0.079	04:31
Air Blank	0.000	04:32
Diagnostics Check	OK	04:32

Data Download Date/Time: 04/06/2017 10:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: CLOSIUS,RICHARD
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 03/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:13

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 03/31/2017 03:39:35

Diagnostics Check	OK	03:33
Air Blank	0.000	03:33
Control Test	0.079	03:34
Air Blank	0.000	03:34
Subject Sample # 1	0.089	03:35
Breath Volume	3.066	03:35
Air Blank	0.000	03:35
Air Blank	0.000	03:37
Subject Sample # 2	0.094	03:38
Breath Volume	2.546	03:38
Air Blank	0.000	03:38
Control Test	0.078	03:38
Air Blank	0.000	03:39
Diagnostics Check	OK	03:39

Data Download Date/Time: 04/06/2017 10:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURGOS,O,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 03/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: PEREZ SANCHEZ

DATE/TIME OF TEST: 03/02/2017 02:44:10

Diagnostics Check	OK	02:36
Air Blank	0.000	02:36
Control Test	0.080	02:36
Air Blank	0.000	02:37
Subject Sample # 1	0.048	02:39
Breath Volume	1.445	02:39
Air Blank	0.000	02:39
Air Blank	0.000	02:41
Subject Sample # 2	0.057	02:42
Breath Volume	2.167	02:42
Air Blank	0.000	02:43
Control Test	0.079	02:43
Air Blank	0.000	02:44
Diagnostics Check	OK	02:44

Data Download Date/Time: 04/14/2017 12:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BONILLA,JOSE,I
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: BONILLA,JOSE
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 03/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 06:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:28

SUBJECT'S LAST NAME: MARTINEZ GARCIGA

DATE/TIME OF TEST: 03/04/2017 06:55:52

Diagnostics Check	OK	06:49
Air Blank	0.000	06:49
Control Test	0.080	06:49
Air Blank	0.000	06:50
Subject Sample # 1	0.181	06:51
Breath Volume	2.730	06:51
Air Blank	0.000	06:52
Air Blank	0.000	06:53
Subject Sample # 2	0.167	06:54
Breath Volume	2.367	06:54
Air Blank	0.000	06:54
Control Test	0.077	06:55
Air Blank	0.000	06:55
Diagnostics Check	OK	06:55

Data Download Date/Time: 04/14/2017 12:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRANCO,GREGORY,R
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: FRANCO,GREGORY
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 03/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:38

SUBJECT'S LAST NAME: LOPEZ-CHAVEZ-VAZQUEZ

DATE/TIME OF TEST: 03/05/2017 05:09:01

Diagnostics Check	OK	05:02
Air Blank	0.000	05:02
Control Test	0.082	05:02
Air Blank	0.000	05:03
Subject Sample # 1	0.205	05:03
Breath Volume	1.671	05:03
Air Blank	0.000	05:04
Air Blank	0.000	05:06
Subject Sample # 2	0.222	05:07
Breath Volume	2.492	05:07
Air Blank	0.000	05:08
Control Test	0.080	05:08
Air Blank	0.000	05:08
Diagnostics Check	OK	05:09

Data Download Date/Time: 04/14/2017 12:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYRTIL,REGINALD,J
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: FRANCO,GREGORY
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 03/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:36

SUBJECT'S LAST NAME: DELOS RIOS RODRIGUEZ

DATE/TIME OF TEST: 03/08/2017 01:04:46

Diagnostics Check		OK	00:58
Air Blank		0.000	00:58
Control Test		0.080	00:59
Air Blank		0.000	00:59
Subject Sample	# 1	0.221	00:59
Breath Volume		1.988	00:59
Air Blank		0.000	01:00
Air Blank		0.000	01:02
Subject Sample	# 2	0.212	01:03
Breath Volume		2.929	01:03
Air Blank		0.000	01:03
Control Test		0.079	01:04
Air Blank		0.000	01:04
Diagnostics Check		OK	01:04

Data Download Date/Time: 04/14/2017 12:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURGOS,O,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: VARGAS,D
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 03/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: OROZCO

DATE/TIME OF TEST: 03/08/2017 02:36:53

Diagnostics Check	OK	02:30
Air Blank	0.000	02:31
Control Test	0.080	02:31
Air Blank	0.000	02:32
Subject Sample # 1	0.159	02:32
Breath Volume	1.710	02:32
Air Blank	0.000	02:33
Air Blank	0.000	02:34
Subject Sample # 2	0.156	02:35
Breath Volume	1.796	02:35
Air Blank	0.000	02:35
Control Test	0.080	02:36
Air Blank	0.000	02:36
Diagnostics Check	OK	02:36

Data Download Date/Time: 04/14/2017 12:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MESA,JONATHAN,A
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: MESA,JONATHAN
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 03/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: LAVERDE

DATE/TIME OF TEST: 03/09/2017 02:27:25

Diagnostics Check	OK	02:19
Air Blank	0.000	02:20
Control Test	0.080	02:20
Air Blank	0.000	02:21
Subject Sample # 1	0.117	02:22
Breath Volume	1.699	02:22
Air Blank	0.000	02:23
Air Blank	0.000	02:25
Subject Sample # 2	0.121	02:25
Breath Volume	1.781	02:25
Air Blank	0.000	02:26
Control Test	0.080	02:26
Air Blank	0.000	02:27
Diagnostics Check	OK	02:27

Data Download Date/Time: 04/14/2017 12:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYRTIL,REGINALD,J
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: FRANCO,GREGORY
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 03/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:28

SUBJECT'S LAST NAME: SANCHEZ

DATE/TIME OF TEST: 03/09/2017 04:47:32

Diagnostics Check		OK	04:35	
Air Blank		0.000	04:36	
Control Test		0.080	04:36	
Air Blank		0.000	04:37	
Subject Sample	# 1	0.135	04:40	Volume Not Met
Breath Volume		0.078	04:40	
Air Blank		0.000	04:41	
Air Blank		0.000	04:42	
Subject Sample	# 2	0.130	04:45	Volume Not Met
Breath Volume		0.031	04:45	
Air Blank		0.000	04:46	
Control Test		0.079	04:46	
Air Blank		0.000	04:47	
Diagnostics Check		OK	04:47	

Data Download Date/Time: 04/14/2017 12:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYRTIL,REGINALD,J
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: FRANCO,GREGORY
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 03/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: SANCHEZ

DATE/TIME OF TEST: 03/09/2017 04:58:39

Diagnostics Check	OK	04:50
Air Blank	0.000	04:51
Control Test	0.079	04:51
Air Blank	0.000	04:52
Subject Sample # 1	0.136	04:54
Breath Volume	1.210	04:54
Air Blank	0.000	04:54
Air Blank	0.000	04:56
Subject Sample # 2	0.133	04:57
Breath Volume	1.285	04:57
Air Blank	0.000	04:57
Control Test	0.078	04:58
Air Blank	0.000	04:58
Diagnostics Check	OK	04:58

Data Download Date/Time: 04/14/2017 12:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VARGAS,D,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: VARGAS,D
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 03/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: MARRERO

DATE/TIME OF TEST: 03/10/2017 05:03:18

Diagnostics Check	OK	04:54	
Air Blank	0.000	04:55	
Control Test	0.080	04:55	
Air Blank	0.000	04:56	
Subject Sample	# 1	04:59	No Sample Provided
Breath Volume	0.000	04:59	
Air Blank	0.000	04:59	
Air Blank	0.000	05:01	
Subject Sample	# 2	05:01	Subject Test Refused
Breath Volume	0.000	05:01	
Air Blank	0.000	05:02	
Control Test	0.080	05:02	
Air Blank	0.000	05:03	
Diagnostics Check	OK	05:03	

Data Download Date/Time: 04/14/2017 12:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 03/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:35

SUBJECT'S LAST NAME: ROA

DATE/TIME OF TEST: 03/10/2017 05:21:03

Diagnostics Check	OK	05:14
Air Blank	0.000	05:14
Control Test	0.079	05:15
Air Blank	0.000	05:15
Subject Sample # 1	0.196	05:16
Breath Volume	2.558	05:16
Air Blank	0.000	05:17
Air Blank	0.000	05:18
Subject Sample # 2	0.197	05:19
Breath Volume	2.949	05:19
Air Blank	0.000	05:20
Control Test	0.079	05:20
Air Blank	0.000	05:20
Diagnostics Check	OK	05:21

Data Download Date/Time: 04/14/2017 12:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 03/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:23

SUBJECT'S LAST NAME: CORDOBA

DATE/TIME OF TEST: 03/11/2017 02:47:28

Diagnostics Check	OK	02:44	
Air Blank	0.000	02:45	
Control Test	0.080	02:45	
Air Blank	0.000	02:45	
Subject Sample	# 1	02:46	Subject Test Refused
Breath Volume	0.000	02:46	
Air Blank	0.000	02:46	
Control Test	0.079	02:46	
Air Blank	0.000	02:47	
Diagnostics Check	OK	02:47	

Data Download Date/Time: 04/14/2017 12:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PACHECO,GEORGE,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: PACHECO,GEORGE
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 03/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 05:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:57

SUBJECT'S LAST NAME: ACEVEDO

DATE/TIME OF TEST: 03/11/2017 06:35:39

Diagnostics Check	OK	06:32	
Air Blank	0.000	06:33	
Control Test	0.081	06:33	
Air Blank	0.000	06:34	
Subject Sample	# 1	06:34	Subject Test Refused
Breath Volume	0.000	06:34	
Air Blank	0.000	06:34	
Control Test	0.079	06:35	
Air Blank	0.000	06:35	
Diagnostics Check	OK	06:35	

Data Download Date/Time: 04/14/2017 12:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARDELAS,MAYLEE,
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: BARDELAS,MAYLE
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 03/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 13:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:10

SUBJECT'S LAST NAME: MORALES

DATE/TIME OF TEST: 03/16/2017 14:38:50

Diagnostics Check	OK	14:30
Air Blank	0.000	14:31
Control Test	0.081	14:31
Air Blank	0.000	14:31
Subject Sample # 1	0.297	14:33
Breath Volume	1.773	14:33
Air Blank	0.000	14:34
Air Blank	0.000	14:36
Subject Sample # 2	0.302	14:37
Breath Volume	2.789	14:37
Air Blank	0.000	14:37
Control Test	0.076	14:38
Air Blank	0.000	14:38
Diagnostics Check	OK	14:38

Data Download Date/Time: 04/14/2017 12:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 03/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:11

SUBJECT'S LAST NAME: RANGEL

DATE/TIME OF TEST: 03/17/2017 04:40:52

Diagnostics Check	OK	04:38	
Air Blank	0.000	04:38	
Control Test	0.080	04:38	
Air Blank	0.000	04:39	
Subject Sample	# 1	04:39	Subject Test Refused
Breath Volume	0.000	04:39	
Air Blank	0.000	04:39	
Control Test	0.079	04:40	
Air Blank	0.000	04:40	
Diagnostics Check	OK	04:40	

Data Download Date/Time: 04/14/2017 12:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINOJOSA,DANIEL,M
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: HINOJOSA,DANIEL
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 03/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:10

SUBJECT'S LAST NAME: ARAUZ

DATE/TIME OF TEST: 03/17/2017 05:34:07

Diagnostics Check	OK	05:31	
Air Blank	0.000	05:31	
Control Test	0.080	05:31	
Air Blank	0.000	05:32	
Subject Sample	# 1	05:32	Subject Test Refused
Breath Volume	0.000	05:32	
Air Blank	0.000	05:33	
Control Test	0.080	05:33	
Air Blank	0.000	05:34	
Diagnostics Check	OK	05:34	

Data Download Date/Time: 04/14/2017 12:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BONILLA,JOSE,I
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: BONILLA,JOSE
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 03/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 05:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:12

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 03/17/2017 05:43:16

Diagnostics Check	OK	05:40	
Air Blank	0.000	05:40	
Control Test	0.080	05:40	
Air Blank	0.000	05:41	
Subject Sample	# 1	05:41	Subject Test Refused
Breath Volume	0.000	05:41	
Air Blank	0.000	05:42	
Control Test	0.080	05:42	
Air Blank	0.000	05:43	
Diagnostics Check	OK	05:43	

Data Download Date/Time: 04/14/2017 12:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BONILLA,JOSE,I
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: BORGES,SEAN
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 03/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 05:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:45

SUBJECT'S LAST NAME: VALDIVIE ROBAINA

DATE/TIME OF TEST: 03/19/2017 06:13:25

Diagnostics Check		OK	06:05
Air Blank		0.000	06:06
Control Test		0.080	06:06
Air Blank		0.000	06:06
Subject Sample	# 1	0.156	06:08
Breath Volume		1.562	06:08
Air Blank		0.000	06:09
Air Blank		0.000	06:11
Subject Sample	# 2	0.168	06:11
Breath Volume		2.085	06:11
Air Blank		0.000	06:12
Control Test		0.077	06:12
Air Blank		0.000	06:13
Diagnostics Check		OK	06:13

Data Download Date/Time: 04/14/2017 12:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALLARDO,JORGE,L
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: GALLARDO,JORGE
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 03/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:05

SUBJECT'S LAST NAME: MARTINEZ URQUIA

DATE/TIME OF TEST: 03/19/2017 21:29:10

Diagnostics Check	OK	21:26	
Air Blank	0.000	21:26	
Control Test	0.080	21:27	
Air Blank	0.000	21:27	
Subject Sample	# 1	21:27	Subject Test Refused
Breath Volume	0.000	21:27	
Air Blank	0.000	21:28	
Control Test	0.080	21:28	
Air Blank	0.000	21:29	
Diagnostics Check	OK	21:29	

Data Download Date/Time: 04/14/2017 12:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: CLOSIUS,RICHARD
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 03/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:08

SUBJECT'S LAST NAME: FLEURY

DATE/TIME OF TEST: 03/24/2017 02:35:55

Diagnostics Check	OK	02:29
Air Blank	0.000	02:30
Control Test	0.081	02:30
Air Blank	0.000	02:30
Subject Sample # 1	0.148	02:31
Breath Volume	3.613	02:31
Air Blank	0.000	02:32
Air Blank	0.000	02:33
Subject Sample # 2	0.155	02:34
Breath Volume	3.417	02:34
Air Blank	0.000	02:35
Control Test	0.080	02:35
Air Blank	0.000	02:35
Diagnostics Check	OK	02:35

Data Download Date/Time: 04/14/2017 12:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VARGAS,D,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: VARGAS,D
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 03/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:21

SUBJECT'S LAST NAME: ACOSTA

DATE/TIME OF TEST: 03/24/2017 03:50:09

Diagnostics Check	OK	03:43
Air Blank	0.000	03:43
Control Test	0.080	03:44
Air Blank	0.000	03:44
Subject Sample # 1	0.151	03:45
Breath Volume	2.515	03:45
Air Blank	0.000	03:46
Air Blank	0.000	03:47
Subject Sample # 2	0.143	03:48
Breath Volume	1.718	03:48
Air Blank	0.000	03:49
Control Test	0.079	03:49
Air Blank	0.000	03:50
Diagnostics Check	OK	03:50

Data Download Date/Time: 04/14/2017 12:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 03/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:33

SUBJECT'S LAST NAME: CAICEDO

DATE/TIME OF TEST: 03/24/2017 05:06:48

Diagnostics Check	OK	04:58
Air Blank	0.000	04:59
Control Test	0.080	04:59
Air Blank	0.000	05:00
Subject Sample # 1	0.199	05:01
Breath Volume	1.511	05:01
Air Blank	0.000	05:02
Air Blank	0.000	05:04
Subject Sample # 2	0.197	05:05
Breath Volume	1.367	05:05
Air Blank	0.000	05:05
Control Test	0.078	05:06
Air Blank	0.000	05:06
Diagnostics Check	OK	05:06

Data Download Date/Time: 04/14/2017 12:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRANCO,GREGORY,R
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: KINSEY-SMITH,F
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 03/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: ALONSO-JIMENEZ

DATE/TIME OF TEST: 03/30/2017 00:42:33

Diagnostics Check	OK	00:39	
Air Blank	0.000	00:40	
Control Test	0.080	00:40	
Air Blank	0.000	00:40	
Subject Sample	# 1	00:41	Subject Test Refused
Breath Volume	0.000	00:41	
Air Blank	0.000	00:41	
Control Test	0.079	00:41	
Air Blank	0.000	00:42	
Diagnostics Check	OK	00:42	

Data Download Date/Time: 04/14/2017 12:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIANCHI,MARIO,D
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: BIANCHI,MARIO
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 03/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:47

SUBJECT'S LAST NAME: SEQUEIRA

DATE/TIME OF TEST: 03/04/2017 03:16:09

Diagnostics Check	OK	03:08
Air Blank	0.000	03:08
Control Test	0.079	03:09
Air Blank	0.000	03:09
Subject Sample # 1	0.179	03:10
Breath Volume	1.679	03:10
Air Blank	0.000	03:10
Air Blank	0.000	03:12
Subject Sample # 2	0.179	03:14
Breath Volume	1.800	03:14
Air Blank	0.000	03:15
Control Test	0.080	03:15
Air Blank	0.000	03:16
Diagnostics Check	OK	03:16

Data Download Date/Time: 04/11/2017 13:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SLIMAK,M,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: SLIMAK,M
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 03/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: MAGUIRE

DATE/TIME OF TEST: 03/06/2017 01:58:35

Diagnostics Check		OK	01:52
Air Blank		0.000	01:52
Control Test		0.079	01:53
Air Blank		0.000	01:53
Subject Sample	# 1	0.217	01:54
Breath Volume		3.367	01:54
Air Blank		0.000	01:54
Air Blank		0.000	01:56
Subject Sample	# 2	0.217	01:56
Breath Volume		2.917	01:56
Air Blank		0.000	01:57
Control Test		0.079	01:57
Air Blank		0.000	01:58
Diagnostics Check		OK	01:58

Data Download Date/Time: 04/11/2017 13:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SLIMAK,M,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: SLIMAK,M
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 03/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: APONTE

DATE/TIME OF TEST: 03/11/2017 00:30:58

Diagnostics Check	OK	00:25	
Air Blank	0.000	00:26	
Control Test	0.078	00:26	
Air Blank	0.000	00:27	
Subject Sample # 1	0.148	00:27	
Breath Volume	2.250	00:27	
Air Blank	0.000	00:28	
Air Blank	0.000	00:29	
		00:30	RFI Detect
Air Blank		00:30	RFI Detect
Air Blank	0.000	00:30	

Data Download Date/Time: 04/11/2017 13:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SLIMAK,M,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: SLIMAK,M
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 03/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: APONTE

DATE/TIME OF TEST: 03/11/2017 00:40:01

Diagnostics Check	OK	00:33
Air Blank	0.000	00:34
Control Test	0.078	00:34
Air Blank	0.000	00:35
Subject Sample # 1	0.149	00:35
Breath Volume	2.367	00:35
Air Blank	0.000	00:36
Air Blank	0.000	00:37
Subject Sample # 2	0.145	00:38
Breath Volume	2.601	00:38
Air Blank	0.000	00:39
Control Test	0.079	00:39
Air Blank	0.000	00:39
Diagnostics Check	OK	00:40

Data Download Date/Time: 04/11/2017 13:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASTILLO,CHRISTIAN,A
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: RODRIGUEZ,MICHELLE
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 03/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 17:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:56

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 03/11/2017 18:25:42

Diagnostics Check	OK	18:17
Air Blank	0.000	18:17
Control Test	0.079	18:17
Air Blank	0.000	18:18
Subject Sample # 1	0.295	18:20
Breath Volume	1.789	18:20
Air Blank	0.000	18:21
Air Blank	0.000	18:22
Subject Sample # 2	0.304	18:23
Breath Volume	1.976	18:23
Air Blank	0.000	18:24
Control Test	0.077	18:25
Air Blank	0.000	18:25
Diagnostics Check	OK	18:25

Data Download Date/Time: 04/11/2017 13:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRANCO,GREGORY,R
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: FRANCO,GREGORY
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 03/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: CLAUDIO

DATE/TIME OF TEST: 03/15/2017 03:50:03

Diagnostics Check	OK	03:46	
Air Blank	0.000	03:46	
Control Test	0.079	03:47	
Air Blank	0.000	03:47	
Subject Sample	# 1	03:48	Subject Test Refused
Breath Volume	0.000	03:48	
Air Blank	0.000	03:49	
Control Test	0.079	03:49	
Air Blank	0.000	03:49	
Diagnostics Check	OK	03:50	

Data Download Date/Time: 04/11/2017 13:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REYNAUD,GEORGE,
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: MARTIN,CHRISTINA
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 03/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:12

SUBJECT'S LAST NAME: CLARK

DATE/TIME OF TEST: 03/22/2017 22:41:56

Diagnostics Check	OK	22:32	
Air Blank	0.000	22:33	
Control Test	0.079	22:33	
Air Blank	0.000	22:34	
Subject Sample	# 1	22:37	No Sample Provided
Breath Volume	0.000	22:37	
Air Blank	0.000	22:37	
Air Blank	0.000	22:39	
Subject Sample	# 2	22:40	Subject Test Refused
Breath Volume	0.000	22:40	
Air Blank	0.000	22:41	
Control Test	0.079	22:41	
Air Blank	0.000	22:41	
Diagnostics Check	OK	22:41	

Data Download Date/Time: 04/11/2017 13:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SLIMAK,M,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: SLIMAK,M
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 03/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: VINDELL

DATE/TIME OF TEST: 03/23/2017 01:33:58

Diagnostics Check	OK	01:26
Air Blank	0.000	01:26
Control Test	0.079	01:27
Air Blank	0.000	01:27
Subject Sample # 1	0.161	01:28
Breath Volume	1.949	01:28
Air Blank	0.000	01:29
Air Blank	0.000	01:31
Subject Sample # 2	0.148	01:32
Breath Volume	2.070	01:32
Air Blank	0.000	01:33
Control Test	0.079	01:33
Air Blank	0.000	01:33
Diagnostics Check	OK	01:33

Data Download Date/Time: 04/11/2017 13:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHIVER,JOE,W
BREATH TEST OPERATOR AGENCY: Florida City PD
ARRESTING OFFICER: SHIVER,JOE
ARRESTING OFFICER AGENCY: Florida City PD
LAST AGENCY INSPECTION: 03/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 15:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:06

SUBJECT'S LAST NAME: NEWELL

DATE/TIME OF TEST: 03/24/2017 17:34:24

Diagnostics Check	OK	17:32	
Air Blank	0.000	17:32	
Control Test	0.079	17:32	
Air Blank	0.000	17:33	
Subject Sample	# 1	17:33	Sequence Aborted
Breath Volume	0.000	17:33	
Air Blank	0.000	17:34	

Data Download Date/Time: 04/11/2017 13:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHIVER,JOE,W
BREATH TEST OPERATOR AGENCY: Florida City PD
ARRESTING OFFICER: SHIVER,JOE
ARRESTING OFFICER AGENCY: Florida City PD
LAST AGENCY INSPECTION: 03/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 15:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:06

SUBJECT'S LAST NAME: NEWELL

DATE/TIME OF TEST: 03/24/2017 17:41:17

Diagnostics Check	OK	17:36	
Air Blank	0.000	17:36	
Control Test	0.079	17:36	
Air Blank	0.000	17:37	
Subject Sample	# 1	17:40	Sequence Aborted
Breath Volume	0.222	17:40	
Air Blank		17:41	Purge Fail

Data Download Date/Time: 04/11/2017 13:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHIVER,JOE,W
BREATH TEST OPERATOR AGENCY: Florida City PD
ARRESTING OFFICER: SHIVER,JOE
ARRESTING OFFICER AGENCY: Florida City PD
LAST AGENCY INSPECTION: 03/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 15:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:06

SUBJECT'S LAST NAME: NEWELL

DATE/TIME OF TEST: 03/24/2017 17:49:51

Diagnostics Check	OK	17:43	
Air Blank	0.000	17:44	
Control Test	0.077	17:44	
Air Blank	0.000	17:44	
Subject Sample # 1	0.140	17:48	Volume Not Met
Breath Volume	0.042	17:48	
Air Blank		17:48	Purge Fail
Air Blank	0.000	17:49	

Data Download Date/Time: 04/11/2017 13:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHIVER,JOE,W
BREATH TEST OPERATOR AGENCY: Florida City PD
ARRESTING OFFICER: SHIVER,JOE
ARRESTING OFFICER AGENCY: Florida City PD
LAST AGENCY INSPECTION: 03/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 15:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:06

SUBJECT'S LAST NAME: NEWELL

DATE/TIME OF TEST: 03/24/2017 18:01:40

Diagnostics Check	OK	17:56	
Air Blank	0.000	17:56	
Control Test	0.076	17:57	
Air Blank	0.000	17:57	
Subject Sample	# 1	18:00	Subject Test Refused
Breath Volume	0.000	18:00	
Air Blank		18:01	Purge Fail
Air Blank	0.000	18:01	

Data Download Date/Time: 04/11/2017 13:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAPANES,R,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: TAPANES,R
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 03/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:13

SUBJECT'S LAST NAME: MEJIA ORELLANA

DATE/TIME OF TEST: 03/25/2017 00:43:32

Diagnostics Check	OK	00:36
Air Blank	0.000	00:36
Control Test	0.079	00:37
Air Blank	0.000	00:37
Subject Sample # 1	0.184	00:38
Breath Volume	1.437	00:38
Air Blank	0.000	00:39
Air Blank	0.000	00:41
Subject Sample # 2	0.170	00:41
Breath Volume	1.183	00:41
Air Blank	0.000	00:42
Control Test	0.077	00:42
Air Blank	0.000	00:43
Diagnostics Check	OK	00:43

Data Download Date/Time: 04/11/2017 13:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASTILLO,ALEX,B
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: CASTILLO,ALEX
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 03/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:19

SUBJECT'S LAST NAME: ARROLIGA

DATE/TIME OF TEST: 03/25/2017 23:43:21

Diagnostics Check	OK	23:39	
Air Blank	0.000	23:40	
Control Test	0.079	23:40	
Air Blank	0.000	23:41	
Subject Sample	# 1 0.134	23:41	
Breath Volume	2.980	23:41	
Air Blank		23:41	RFI Detect
Air Blank	0.000	23:42	
Control Test	0.080	23:42	
Air Blank	0.000	23:43	
Diagnostics Check	OK	23:43	

Data Download Date/Time: 04/11/2017 13:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASTILLO,ALEX,B
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: CASTILLO,ALEX
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 03/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:19

SUBJECT'S LAST NAME: ARROLIGA

DATE/TIME OF TEST: 03/25/2017 23:50:53

Diagnostics Check	OK	23:44
Air Blank	0.000	23:45
Control Test	0.079	23:45
Air Blank	0.000	23:46
Subject Sample # 1	0.136	23:46
Breath Volume	2.894	23:46
Air Blank	0.000	23:47
Air Blank	0.000	23:48
Subject Sample # 2	0.126	23:49
Breath Volume	2.152	23:49
Air Blank	0.000	23:49
Control Test	0.078	23:50
Air Blank	0.000	23:50
Diagnostics Check	OK	23:50

Data Download Date/Time: 04/11/2017 13:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SLIMAK,M,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: SLIMAK,M
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 03/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: GINORIS

DATE/TIME OF TEST: 03/30/2017 04:21:27

Diagnostics Check	OK	04:18	
Air Blank	0.000	04:19	
Control Test	0.079	04:19	
Air Blank	0.000	04:19	
Subject Sample	# 1	04:20	Subject Test Refused
Breath Volume	0.000	04:20	
Air Blank	0.000	04:20	
Control Test	0.079	04:20	
Air Blank	0.000	04:21	
Diagnostics Check	OK	04:21	

Data Download Date/Time: 04/11/2017 13:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZEVEDO,MIKE,C
BREATH TEST OPERATOR AGENCY: North Miami Beach PD
ARRESTING OFFICER: OLIVARES,ANDRES
ARRESTING OFFICER AGENCY: North Miami Beach PD
LAST AGENCY INSPECTION: 02/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:01

SUBJECT'S LAST NAME: DUBON

DATE/TIME OF TEST: 03/02/2017 00:31:03

Diagnostics Check	OK	00:24	
Air Blank	0.000	00:25	
Control Test	0.077	00:25	
Air Blank	0.000	00:25	
Subject Sample	# 1 0.160	00:27	
Breath Volume	1.167	00:27	
Air Blank	0.000	00:27	
Air Blank	0.000	00:29	
Subject Sample	# 2	00:29	RFI Detect
Breath Volume	0.000	00:29	
Air Blank	0.000	00:30	
Control Test	0.078	00:30	
Air Blank	0.000	00:30	
Diagnostics Check	OK	00:31	

Data Download Date/Time: 03/30/2017 23:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZEVEDO,MIKE,C
BREATH TEST OPERATOR AGENCY: North Miami Beach PD
ARRESTING OFFICER: OLIVARES,ANDRES
ARRESTING OFFICER AGENCY: North Miami Beach PD
LAST AGENCY INSPECTION: 02/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:01

SUBJECT'S LAST NAME: DUBON

DATE/TIME OF TEST: 03/02/2017 00:41:03

Diagnostics Check	OK	00:32	
Air Blank	0.000	00:32	
Control Test	0.078	00:33	
Air Blank	0.000	00:33	
Subject Sample	# 1	00:36	No Sample Provided
Breath Volume	0.000	00:36	
Air Blank	0.000	00:37	
Air Blank	0.000	00:39	
Subject Sample	# 2	00:39	Subject Test Refused
Breath Volume	0.000	00:39	
Air Blank	0.000	00:40	
Control Test	0.078	00:40	
Air Blank	0.000	00:40	
Diagnostics Check	OK	00:41	

Data Download Date/Time: 03/30/2017 23:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OLIVARES,ANDRES,J
BREATH TEST OPERATOR AGENCY: North Miami Beach PD
ARRESTING OFFICER: AZEVEDO,MIKE
ARRESTING OFFICER AGENCY: North Miami Beach PD
LAST AGENCY INSPECTION: 02/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:47

SUBJECT'S LAST NAME: BILLIE

DATE/TIME OF TEST: 03/03/2017 02:09:05

Diagnostics Check	OK	02:07	
Air Blank	0.000	02:07	
Control Test	0.078	02:07	
Air Blank	0.000	02:08	
Subject Sample	# 1	02:08	RFI Detect
Breath Volume	0.000	02:08	
Air Blank		02:08	RFI Detect
Air Blank		02:09	RFI Detect

Data Download Date/Time: 03/30/2017 23:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OLIVARES,ANDRES,J
BREATH TEST OPERATOR AGENCY: North Miami Beach PD
ARRESTING OFFICER: AZEVEDO,MIKE
ARRESTING OFFICER AGENCY: North Miami Beach PD
LAST AGENCY INSPECTION: 02/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:47

SUBJECT'S LAST NAME: BILLIE

DATE/TIME OF TEST: 03/03/2017 02:17:33

Diagnostics Check	OK	02:12	
Air Blank	0.000	02:12	
Control Test	0.078	02:12	
Air Blank	0.000	02:13	
Subject Sample	# 1 0.191	02:13	
Breath Volume	1.578	02:13	
Air Blank	0.000	02:14	
Air Blank	0.000	02:16	
Subject Sample	# 2 0.174	02:16	
Breath Volume	1.777	02:16	
Air Blank		02:17	RFI Detect
Air Blank	0.000	02:17	

Data Download Date/Time: 03/30/2017 23:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OLIVARES,ANDRES,J
BREATH TEST OPERATOR AGENCY: North Miami Beach PD
ARRESTING OFFICER: AZEVEDO,MIKE
ARRESTING OFFICER AGENCY: North Miami Beach PD
LAST AGENCY INSPECTION: 02/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:47

SUBJECT'S LAST NAME: BILLIE

DATE/TIME OF TEST: 03/03/2017 02:26:21

Diagnostics Check	OK	02:20
Air Blank	0.000	02:20
Control Test	0.077	02:20
Air Blank	0.000	02:21
Subject Sample # 1	0.177	02:21
Breath Volume	1.992	02:21
Air Blank	0.000	02:22
Air Blank	0.000	02:24
Subject Sample # 2	0.173	02:24
Breath Volume	2.343	02:24
Air Blank	0.000	02:25
Control Test	0.076	02:25
Air Blank	0.000	02:26
Diagnostics Check	OK	02:26

Data Download Date/Time: 03/30/2017 23:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OLIVARES,AJ,
BREATH TEST OPERATOR AGENCY: North Miami Beach PD
ARRESTING OFFICER: AZEVEDO,MIKE
ARRESTING OFFICER AGENCY: North Miami Beach PD
LAST AGENCY INSPECTION: 02/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:35

SUBJECT'S LAST NAME: JUSTINIEN

DATE/TIME OF TEST: 03/03/2017 05:01:47

Diagnostics Check	OK	04:55
Air Blank	0.000	04:56
Control Test	0.079	04:56
Air Blank	0.000	04:56
Subject Sample # 1	0.000	04:57
Breath Volume	2.242	04:57
Air Blank	0.000	04:58
Air Blank	0.000	05:00
Subject Sample # 2	0.000	05:00
Breath Volume	2.648	05:00
Air Blank	0.000	05:00
Control Test	0.077	05:01
Air Blank	0.000	05:01
Diagnostics Check	OK	05:01

Data Download Date/Time: 03/30/2017 23:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OLIVARES,AJ,
BREATH TEST OPERATOR AGENCY: North Miami Beach PD
ARRESTING OFFICER: OLIVARES,AJ
ARRESTING OFFICER AGENCY: North Miami Beach PD
LAST AGENCY INSPECTION: 02/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:08

SUBJECT'S LAST NAME: FARQUHARSON

DATE/TIME OF TEST: 03/04/2017 00:44:31

Diagnostics Check	OK	00:28	
Air Blank	0.000	00:29	
Control Test	0.078	00:29	
Air Blank	0.000	00:30	
Subject Sample	# 1 0.236	00:31	
Breath Volume	1.121	00:31	
Air Blank	0.000	00:32	
Air Blank	0.000	00:34	
Subject Sample	# 2 0.210	00:37	Volume Not Met
Breath Volume	0.601	00:37	
Air Blank	0.000	00:38	
Air Blank	0.000	00:39	
Subject Sample	# 3 0.232	00:42	Volume Not Met
Breath Volume	0.144	00:42	
Air Blank	0.000	00:43	
Control Test	0.074	00:43	Control Outside Tolerance
Air Blank	0.000	00:44	

Data Download Date/Time: 03/30/2017 23:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OLIVARES,AJ,
BREATH TEST OPERATOR AGENCY: North Miami Beach PD
ARRESTING OFFICER: AZEVEDO,MIKE
ARRESTING OFFICER AGENCY: North Miami Beach PD
LAST AGENCY INSPECTION: 02/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:09

SUBJECT'S LAST NAME: BY

DATE/TIME OF TEST: 03/05/2017 01:40:32

Diagnostics Check	OK	01:29	
Air Blank	0.000	01:29	
Control Test	0.079	01:29	
Air Blank	0.000	01:30	
Subject Sample	# 1	01:33	No Sample Provided
Breath Volume	0.000	01:33	
Air Blank	0.000	01:34	
Air Blank	0.000	01:36	
Subject Sample	# 2	01:39	No Sample Provided
Breath Volume	0.000	01:39	
Air Blank	0.000	01:39	
Control Test	0.079	01:39	
Air Blank	0.000	01:40	
Diagnostics Check	OK	01:40	

Data Download Date/Time: 03/30/2017 23:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZEVEDO,MIKE,C
BREATH TEST OPERATOR AGENCY: North Miami Beach PD
ARRESTING OFFICER: KELLNER,DOUGLAS
ARRESTING OFFICER AGENCY: North Miami Beach PD
LAST AGENCY INSPECTION: 02/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:54

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 03/10/2017 00:17:25

Diagnostics Check	OK	00:14	
Air Blank	0.000	00:15	
Control Test	0.078	00:15	
Air Blank	0.000	00:15	
Subject Sample	# 1	00:16	Subject Test Refused
Breath Volume	0.000	00:16	
Air Blank	0.000	00:16	
Control Test	0.079	00:16	
Air Blank	0.000	00:17	
Diagnostics Check	OK	00:17	

Data Download Date/Time: 03/30/2017 23:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZEVEDO,MIKE,C
BREATH TEST OPERATOR AGENCY: North Miami Beach PD
ARRESTING OFFICER: FARINI,KAIO
ARRESTING OFFICER AGENCY: North Miami Beach PD
LAST AGENCY INSPECTION: 02/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:41

SUBJECT'S LAST NAME: HARTLEY

DATE/TIME OF TEST: 03/10/2017 02:08:33

Diagnostics Check	OK	02:01
Air Blank	0.000	02:02
Control Test	0.078	02:02
Air Blank	0.000	02:02
Subject Sample # 1	0.141	02:03
Breath Volume	1.664	02:03
Air Blank	0.000	02:04
Air Blank	0.000	02:06
Subject Sample # 2	0.138	02:06
Breath Volume	1.976	02:06
Air Blank	0.000	02:07
Control Test	0.078	02:07
Air Blank	0.000	02:08
Diagnostics Check	OK	02:08

Data Download Date/Time: 03/30/2017 23:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZEVEDO,MIKE,C
BREATH TEST OPERATOR AGENCY: North Miami Beach PD
ARRESTING OFFICER: AZEVEDO,MIKE
ARRESTING OFFICER AGENCY: North Miami Beach PD
LAST AGENCY INSPECTION: 02/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:27

SUBJECT'S LAST NAME: MICELI

DATE/TIME OF TEST: 03/12/2017 00:54:15

Diagnostics Check	OK	00:47
Air Blank	0.000	00:48
Control Test	0.078	00:48
Air Blank	0.000	00:48
Subject Sample # 1	0.189	00:49
Breath Volume	2.324	00:49
Air Blank	0.000	00:50
Air Blank	0.000	00:52
Subject Sample # 2	0.196	00:52
Breath Volume	1.605	00:52
Air Blank	0.000	00:53
Control Test	0.078	00:53
Air Blank	0.000	00:54
Diagnostics Check	OK	00:54

Data Download Date/Time: 03/30/2017 23:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OLIVARES,AJ,
BREATH TEST OPERATOR AGENCY: North Miami Beach PD
ARRESTING OFFICER: AZEVEDO,MIKE
ARRESTING OFFICER AGENCY: North Miami Beach PD
LAST AGENCY INSPECTION: 02/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:38

SUBJECT'S LAST NAME: BENJAMIN

DATE/TIME OF TEST: 03/18/2017 02:05:42

Diagnostics Check	OK	01:58
Air Blank	0.000	01:59
Control Test	0.079	01:59
Air Blank	0.000	01:59
Subject Sample # 1	0.077	02:01
Breath Volume	1.406	02:01
Air Blank	0.000	02:01
Air Blank	0.000	02:03
Subject Sample # 2	0.080	02:04
Breath Volume	1.621	02:04
Air Blank	0.000	02:04
Control Test	0.078	02:05
Air Blank	0.000	02:05
Diagnostics Check	OK	02:05

Data Download Date/Time: 03/30/2017 23:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHAPIRO,SABINA,M
BREATH TEST OPERATOR AGENCY: Aventura PD
ARRESTING OFFICER: KIRKUS,JOSHUA
ARRESTING OFFICER AGENCY: North Miami Beach PD
LAST AGENCY INSPECTION: 02/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 15:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:45

SUBJECT'S LAST NAME: KUPCZYK

DATE/TIME OF TEST: 03/19/2017 16:12:44

Diagnostics Check	OK	16:05	
Air Blank	0.000	16:06	
Control Test	0.078	16:06	
Air Blank	0.000	16:07	
Subject Sample # 1	0.167	16:08	
Breath Volume	1.269	16:08	
Air Blank	0.000	16:08	
Air Blank	0.000	16:10	
Subject Sample # 2	0.167	16:11	
Breath Volume	2.113	16:11	
Air Blank	0.000	16:11	
Control Test	0.074	16:12	Control Outside Tolerance
Air Blank	0.000	16:12	

Data Download Date/Time: 03/30/2017 23:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AJO,MICHAEL,D
BREATH TEST OPERATOR AGENCY: North Miami Beach PD
ARRESTING OFFICER: AJO,MICHAEL
ARRESTING OFFICER AGENCY: North Miami Beach PD
LAST AGENCY INSPECTION: 02/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: AUPONT

DATE/TIME OF TEST: 03/20/2017 00:30:20

Diagnostics Check	OK	00:21
Air Blank	0.000	00:22
Control Test	0.078	00:22
Air Blank	0.000	00:22
Subject Sample # 1	0.155	00:25
Breath Volume	1.269	00:25
Air Blank	0.000	00:26
Air Blank	0.000	00:27
Subject Sample # 2	0.162	00:28
Breath Volume	1.812	00:28
Air Blank	0.000	00:29
Control Test	0.077	00:29
Air Blank	0.000	00:30
Diagnostics Check	OK	00:30

Data Download Date/Time: 03/30/2017 23:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AJO,MICHAEL,D
BREATH TEST OPERATOR AGENCY: North Miami Beach PD
ARRESTING OFFICER: TANIS,MIKE
ARRESTING OFFICER AGENCY: North Miami Beach PD
LAST AGENCY INSPECTION: 02/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:04

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 03/29/2017 01:28:07

Diagnostics Check	OK	01:25	
Air Blank	0.000	01:25	
Control Test	0.078	01:25	
Air Blank	0.000	01:26	
Subject Sample	# 1	01:26	Subject Test Refused
Breath Volume	0.000	01:26	
Air Blank	0.000	01:27	
Control Test	0.078	01:27	
Air Blank	0.000	01:28	
Diagnostics Check	OK	01:28	

Data Download Date/Time: 03/30/2017 23:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZEVEDO,MIKE,C
BREATH TEST OPERATOR AGENCY: North Miami Beach PD
ARRESTING OFFICER: AZEVEDO,MIKE
ARRESTING OFFICER AGENCY: North Miami Beach PD
LAST AGENCY INSPECTION: 03/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:07

SUBJECT'S LAST NAME: TRUJILLO ZORRILLA

DATE/TIME OF TEST: 03/30/2017 23:36:03

Diagnostics Check	OK	23:27	
Air Blank	0.000	23:27	
Control Test	0.078	23:28	
Air Blank	0.000	23:28	
Subject Sample	# 1	23:31	No Sample Provided
Breath Volume	0.000	23:31	
Air Blank	0.000	23:32	
Air Blank	0.000	23:34	
Subject Sample	# 2	23:34	Subject Test Refused
Breath Volume	0.000	23:34	
Air Blank	0.000	23:35	
Control Test	0.079	23:35	
Air Blank	0.000	23:35	
Diagnostics Check	OK	23:36	

Data Download Date/Time: 04/26/2017 19:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000887

INSTRUMENT REGISTERED TO: CHATTAHOOCHEE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POSTON,WILLIAM,A
BREATH TEST OPERATOR AGENCY: Chattahoochee PD
ARRESTING OFFICER: BUSH,TYLER
ARRESTING OFFICER AGENCY: Chattahoochee PD
LAST AGENCY INSPECTION: 02/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A4
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 18:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:20

SUBJECT'S LAST NAME: MITCHELL

DATE/TIME OF TEST: 03/05/2017 18:55:33

Diagnostics Check	OK	18:53	
Air Blank	0.000	18:53	
Control Test	0.078	18:53	
Air Blank	0.000	18:54	
Subject Sample	# 1	18:55	Sequence Aborted
Breath Volume	0.000	18:55	
Air Blank	0.000	18:55	

Data Download Date/Time: 07/10/2017 23:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000906

INSTRUMENT REGISTERED TO: FISH & WILDLIFE CC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PLUSSA,KYLE,A
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: PLUSSA,KYLE
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 03/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: BUI
ARREST TIME: 19:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:25

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 03/25/2017 20:01:18

Diagnostics Check	OK	19:55
Air Blank	0.000	19:55
Control Test	0.079	19:56
Air Blank	0.000	19:56
Subject Sample # 1	0.019	19:57
Breath Volume	1.359	19:57
Air Blank	0.000	19:57
Air Blank	0.000	19:59
Subject Sample # 2	0.019	19:59
Breath Volume	2.683	19:59
Air Blank	0.000	20:00
Control Test	0.081	20:00
Air Blank	0.000	20:01
Diagnostics Check	OK	20:01

Data Download Date/Time: 04/13/2017 16:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLAFLIN,DAVID,S
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: SMITH,JOSIAH
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 03/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:17

SUBJECT'S LAST NAME: WEEDALL

DATE/TIME OF TEST: 03/31/2017 22:51:54

Diagnostics Check	OK	22:45	
Air Blank	0.000	22:45	
Control Test	0.076	22:46	
Air Blank	0.000	22:46	
Subject Sample	# 1 0.000	22:47	
Breath Volume	3.027	22:47	
Air Blank	0.000	22:47	
Air Blank	0.000	22:49	
Subject Sample	# 2 0.000	22:50	
Breath Volume	2.875	22:50	
Air Blank	0.000	22:50	
Control Test	0.080	22:50	
Air Blank	0.000	22:51	
Diagnostics Check	Fail	22:51	Analytical Stability Test
Air Blank	0.000	22:51	

Data Download Date/Time: 04/12/2017 11:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLAFLIN,DAVID,S
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: SMITH,JOSIAH
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 03/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:17

SUBJECT'S LAST NAME: WEEDALL

DATE/TIME OF TEST: 03/31/2017 23:01:32

Diagnostics Check	OK	22:55	
Air Blank	0.000	22:55	
Control Test	0.077	22:56	
Air Blank	0.000	22:56	
Subject Sample # 1	0.000	22:57	
Breath Volume	3.125	22:57	
Air Blank	0.000	22:57	
Air Blank	0.000	22:59	
Subject Sample # 2	0.007	23:00	
Breath Volume	2.847	23:00	
Air Blank	0.000	23:00	
Control Test	0.083	23:01	Interferent Detect
Air Blank	0.000	23:01	

Data Download Date/Time: 04/12/2017 11:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,JOSIAH,R
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: SMITH,JOSIAH
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 02/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23016080A3
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:53

SUBJECT'S LAST NAME: MANTIA

DATE/TIME OF TEST: 03/01/2017 01:23:27

Diagnostics Check	OK	01:15
Air Blank	0.000	01:15
Control Test	0.081	01:16
Air Blank	0.000	01:16
Subject Sample # 1	0.000	01:18
Breath Volume	1.812	01:18
Air Blank	0.000	01:18
Air Blank	0.000	01:20
Subject Sample # 2	0.000	01:22
Breath Volume	1.703	01:22
Air Blank	0.000	01:22
Control Test	0.079	01:22
Air Blank	0.000	01:23
Diagnostics Check	OK	01:23

Data Download Date/Time: 03/08/2017 11:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALGER, MICHAEL, C
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: ALGER, MICHAEL
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 02/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23016080A3
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:56

SUBJECT'S LAST NAME: HOTCHKISS

DATE/TIME OF TEST: 03/01/2017 02:24:40

Diagnostics Check	OK	02:18
Air Blank	0.000	02:18
Control Test	0.080	02:18
Air Blank	0.000	02:19
Subject Sample # 1	0.165	02:20
Breath Volume	2.457	02:20
Air Blank	0.000	02:20
Air Blank	0.000	02:22
Subject Sample # 2	0.156	02:23
Breath Volume	1.562	02:23
Air Blank	0.000	02:23
Control Test	0.080	02:24
Air Blank	0.000	02:24
Diagnostics Check	OK	02:24

Data Download Date/Time: 03/08/2017 11:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,JOSIAH,R
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: SMITH,JOSIAH
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 02/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23016080A3
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:26

SUBJECT'S LAST NAME: LEIVA

DATE/TIME OF TEST: 03/04/2017 03:52:58

Diagnostics Check		OK	03:46
Air Blank		0.000	03:46
Control Test		0.080	03:47
Air Blank		0.000	03:47
Subject Sample	# 1	0.185	03:48
Breath Volume		3.003	03:48
Air Blank		0.000	03:49
Air Blank		0.000	03:50
Subject Sample	# 2	0.182	03:51
Breath Volume		2.394	03:51
Air Blank		0.000	03:52
Control Test		0.080	03:52
Air Blank		0.000	03:52
Diagnostics Check		OK	03:52

Data Download Date/Time: 03/08/2017 11:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH, JOSIAH, R
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: CALHOUN, K
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 02/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23016080A3
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:24

SUBJECT'S LAST NAME: ARNE

DATE/TIME OF TEST: 03/04/2017 04:05:29

Diagnostics Check	OK	03:57
Air Blank	0.000	03:57
Control Test	0.080	03:57
Air Blank	0.000	03:58
Subject Sample # 1	0.100	04:00
Breath Volume	2.425	04:00
Air Blank	0.000	04:01
Air Blank	0.000	04:02
Subject Sample # 2	0.094	04:03
Breath Volume	2.398	04:03
Air Blank	0.000	04:04
Control Test	0.079	04:04
Air Blank	0.000	04:05
Diagnostics Check	OK	04:05

Data Download Date/Time: 03/08/2017 11:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,JOSIAH,R
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: SMITH,JOSIAH
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 02/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23016080A3
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:48

SUBJECT'S LAST NAME: ;6360101253600388502

DATE/TIME OF TEST: 03/05/2017 00:15:31

Diagnostics Check	OK	00:08
Air Blank	0.000	00:09
Control Test	0.080	00:09
Air Blank	0.000	00:10
Subject Sample # 1	0.145	00:11
Breath Volume	2.664	00:11
Air Blank	0.000	00:11
Air Blank	0.000	00:13
Subject Sample # 2	0.149	00:13
Breath Volume	1.648	00:13
Air Blank	0.000	00:14
Control Test	0.079	00:14
Air Blank	0.000	00:15
Diagnostics Check	OK	00:15

Data Download Date/Time: 03/08/2017 11:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDIN, JEREMY, L
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: TOTU, ROBERT
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 02/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23016080A3
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:35

SUBJECT'S LAST NAME: JERONIMO HERNANDEZ

DATE/TIME OF TEST: 03/05/2017 05:33:41

Diagnostics Check	OK	05:27
Air Blank	0.000	05:28
Control Test	0.079	05:28
Air Blank	0.000	05:28
Subject Sample # 1	0.198	05:29
Breath Volume	1.492	05:29
Air Blank	0.000	05:29
Air Blank	0.000	05:31
Subject Sample # 2	0.206	05:32
Breath Volume	2.625	05:32
Air Blank	0.000	05:32
Control Test	0.079	05:33
Air Blank	0.000	05:33
Diagnostics Check	OK	05:33

Data Download Date/Time: 03/08/2017 11:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUNCAN,N,
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: DUNCAN,N
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 02/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23016080A3
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:31

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 03/07/2017 00:02:33

Diagnostics Check	OK	23:54
Air Blank	0.000	23:55
Control Test	0.080	23:55
Air Blank	0.000	23:55
Subject Sample # 2	0.248	23:56
Breath Volume	1.328	23:56
Air Blank	0.000	23:56
Air Blank	0.000	23:58
Subject Sample # 1	0.240	00:00
Breath Volume	1.421	00:00
Air Blank	0.000	00:01
Control Test	0.081	00:01
Air Blank	0.000	00:02
Diagnostics Check	OK	00:02

Data Download Date/Time: 03/08/2017 11:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUNCAN,N,
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: DUNCAN,N
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 02/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23016080A3
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:52

SUBJECT'S LAST NAME: JEAN

DATE/TIME OF TEST: 03/08/2017 05:23:11

Diagnostics Check	OK	05:16
Air Blank	0.000	05:17
Control Test	0.080	05:17
Air Blank	0.000	05:18
Subject Sample # 1	0.000	05:18
Breath Volume	2.058	05:18
Air Blank	0.000	05:19
Air Blank	0.000	05:21
Subject Sample # 2	0.000	05:21
Breath Volume	1.519	05:21
Air Blank	0.000	05:22
Control Test	0.079	05:22
Air Blank	0.000	05:23
Diagnostics Check	OK	05:23

Data Download Date/Time: 03/08/2017 11:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH, JOSIAH, R
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: SMITH, JOSIAH
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 03/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23016080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 17:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:00

SUBJECT'S LAST NAME: POTTMAN

DATE/TIME OF TEST: 03/08/2017 18:26:56

Diagnostics Check	OK	18:20
Air Blank	0.000	18:21
Control Test	0.080	18:21
Air Blank	0.000	18:22
Subject Sample # 1	0.000	18:22
Breath Volume	2.113	18:22
Air Blank	0.000	18:23
Air Blank	0.000	18:25
Subject Sample # 2	0.000	18:25
Breath Volume	2.171	18:25
Air Blank	0.000	18:26
Control Test	0.081	18:26
Air Blank	0.000	18:26
Diagnostics Check	OK	18:26

Data Download Date/Time: 04/12/2017 11:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUNCAN,N,
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: SAMMUT,A
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 03/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23016080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:16

SUBJECT'S LAST NAME: LEE II

DATE/TIME OF TEST: 03/09/2017 03:38:36

Diagnostics Check	OK	03:32
Air Blank	0.000	03:33
Control Test	0.080	03:33
Air Blank	0.000	03:34
Subject Sample # 1	0.000	03:34
Breath Volume	1.957	03:34
Air Blank	0.000	03:34
Air Blank	0.000	03:36
Subject Sample # 2	0.000	03:37
Breath Volume	1.867	03:37
Air Blank	0.000	03:37
Control Test	0.081	03:38
Air Blank	0.000	03:38
Diagnostics Check	OK	03:38

Data Download Date/Time: 04/12/2017 11:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUNCAN,N,
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: BAIR,J
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 03/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23016080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: ARANGO PEREZ

DATE/TIME OF TEST: 03/09/2017 04:31:55

Diagnostics Check	OK	04:24
Air Blank	0.000	04:24
Control Test	0.080	04:25
Air Blank	0.000	04:25
Subject Sample # 1	0.110	04:26
Breath Volume	2.074	04:26
Air Blank	0.000	04:27
Air Blank	0.000	04:29
Subject Sample # 2	0.104	04:30
Breath Volume	1.828	04:30
Air Blank	0.000	04:30
Control Test	0.077	04:31
Air Blank	0.000	04:31
Diagnostics Check	OK	04:31

Data Download Date/Time: 04/12/2017 11:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRASK,J,
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: TRASK,J
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 03/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23016080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 16:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:20

SUBJECT'S LAST NAME: RONQUILLO-TINOCO

DATE/TIME OF TEST: 03/10/2017 17:45:44

Diagnostics Check	OK	17:40	
Air Blank	0.000	17:41	
Control Test	0.080	17:41	
Air Blank	0.000	17:42	
Subject Sample # 1	0.255	17:44	
Breath Volume	1.226	17:44	
Air Blank		17:45	Purge Fail
Air Blank	0.000	17:45	

Data Download Date/Time: 04/12/2017 11:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRASK,J,
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: TRASK,J
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 03/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23016080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 16:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:20

SUBJECT'S LAST NAME: RONQUILLO-TINOCO

DATE/TIME OF TEST: 03/10/2017 17:53:47

Diagnostics Check	OK	17:47
Air Blank	0.000	17:47
Control Test	0.079	17:47
Air Blank	0.000	17:48
Subject Sample # 1	0.255	17:49
Breath Volume	1.625	17:49
Air Blank	0.000	17:49
Air Blank	0.000	17:51
Subject Sample # 2	0.259	17:52
Breath Volume	1.355	17:52
Air Blank	0.000	17:52
Control Test	0.079	17:53
Air Blank	0.000	17:53
Diagnostics Check	OK	17:53

Data Download Date/Time: 04/12/2017 11:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDIN, JEREMY, L
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: SABOE, EVAN
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 03/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23016080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 11:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:33

SUBJECT'S LAST NAME: LABITZKE

DATE/TIME OF TEST: 03/14/2017 12:05:21

Diagnostics Check	OK	11:59
Air Blank	0.000	12:00
Control Test	0.080	12:00
Air Blank	0.000	12:00
Subject Sample # 1	0.000	12:01
Breath Volume	2.082	12:01
Air Blank	0.000	12:01
Air Blank	0.000	12:03
Subject Sample # 2	0.000	12:03
Breath Volume	1.968	12:03
Air Blank	0.000	12:04
Control Test	0.078	12:04
Air Blank	0.000	12:05
Diagnostics Check	OK	12:05

Data Download Date/Time: 04/12/2017 11:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOULD,LEONARD,
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: NODARSE,YESSENIA
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 03/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23016080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 18:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:15

SUBJECT'S LAST NAME: CAUDILL

DATE/TIME OF TEST: 03/14/2017 20:07:27

Diagnostics Check	OK	20:01
Air Blank	0.000	20:01
Control Test	0.080	20:02
Air Blank	0.000	20:02
Subject Sample # 1	0.134	20:03
Breath Volume	1.800	20:03
Air Blank	0.000	20:03
Air Blank	0.000	20:05
Subject Sample # 2	0.125	20:05
Breath Volume	1.855	20:05
Air Blank	0.000	20:06
Control Test	0.080	20:06
Air Blank	0.000	20:07
Diagnostics Check	OK	20:07

Data Download Date/Time: 04/12/2017 11:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALGER, MICHAEL, C
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: BAIR, JAMES
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 03/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23016080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: PIERCE

DATE/TIME OF TEST: 03/15/2017 04:46:03

Diagnostics Check	OK	04:39
Air Blank	0.000	04:40
Control Test	0.081	04:40
Air Blank	0.000	04:40
Subject Sample # 1	0.000	04:41
Breath Volume	1.371	04:41
Air Blank	0.000	04:42
Air Blank	0.000	04:44
Subject Sample # 2	0.000	04:44
Breath Volume	2.027	04:44
Air Blank	0.000	04:45
Control Test	0.080	04:45
Air Blank	0.000	04:45
Diagnostics Check	OK	04:46

Data Download Date/Time: 04/12/2017 11:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUNCAN,N,
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: GALLANT,K
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 03/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23016080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:32

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 03/16/2017 04:02:14

Diagnostics Check	OK	03:55
Air Blank	0.000	03:56
Control Test	0.081	03:56
Air Blank	0.000	03:57
Subject Sample # 1	0.146	03:57
Breath Volume	1.121	03:57
Air Blank	0.000	03:58
Air Blank	0.000	03:59
Subject Sample # 2	0.151	04:00
Breath Volume	1.515	04:00
Air Blank	0.000	04:01
Control Test	0.080	04:01
Air Blank	0.000	04:02
Diagnostics Check	OK	04:02

Data Download Date/Time: 04/12/2017 11:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENEMAN,LESA,
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: BRENEMAN,LESA
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 03/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23016080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:29

SUBJECT'S LAST NAME: HAN

DATE/TIME OF TEST: 03/17/2017 21:57:09

Diagnostics Check	OK	21:50
Air Blank	0.000	21:51
Control Test	0.081	21:51
Air Blank	0.000	21:51
Subject Sample # 1	0.165	21:52
Breath Volume	2.527	21:52
Air Blank	0.000	21:53
Air Blank	0.000	21:54
Subject Sample # 2	0.166	21:55
Breath Volume	2.703	21:55
Air Blank	0.000	21:56
Control Test	0.080	21:56
Air Blank	0.000	21:57
Diagnostics Check	OK	21:57

Data Download Date/Time: 04/12/2017 11:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NAJARRO,CARLOS,O
BREATH TEST OPERATOR AGENCY: Lee Port Authority
ARRESTING OFFICER: FENSKE,ALEX
ARRESTING OFFICER AGENCY: Lee Port Authority
LAST AGENCY INSPECTION: 03/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23016080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: IRWIN

DATE/TIME OF TEST: 03/17/2017 23:56:40

Diagnostics Check	OK	23:50
Air Blank	0.000	23:51
Control Test	0.080	23:51
Air Blank	0.000	23:52
Subject Sample # 1	0.000	23:52
Breath Volume	2.492	23:52
Air Blank	0.000	23:52
Air Blank	0.000	23:54
Subject Sample # 2	0.000	23:55
Breath Volume	4.433	23:55
Air Blank	0.000	23:55
Control Test	0.081	23:56
Air Blank	0.000	23:56
Diagnostics Check	OK	23:56

Data Download Date/Time: 04/12/2017 11:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRASK,JUSTIN,N
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: TRASK,J
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 03/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23016080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: GUILBE

DATE/TIME OF TEST: 03/18/2017 00:56:32

Diagnostics Check	OK	00:55	
Air Blank	0.000	00:55	
Control Test	0.080	00:55	
Air Blank		00:56	RFI Detect
Air Blank	0.000	00:56	

Data Download Date/Time: 04/12/2017 11:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRASK,J,
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: TRASK,J
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 03/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23016080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: GUILBE

DATE/TIME OF TEST: 03/18/2017 01:06:04

Diagnostics Check	OK	00:57
Air Blank	0.000	00:57
Control Test	0.079	00:58
Air Blank	0.000	00:58
Subject Sample # 1	0.143	01:01
Breath Volume	1.687	01:01
Air Blank	0.000	01:02
Air Blank	0.000	01:03
Subject Sample # 2	0.137	01:04
Breath Volume	1.648	01:04
Air Blank	0.000	01:05
Control Test	0.079	01:05
Air Blank	0.000	01:05
Diagnostics Check	OK	01:06

Data Download Date/Time: 04/12/2017 11:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NAJARRO,CARLOS,O
BREATH TEST OPERATOR AGENCY: Lee Port Authority
ARRESTING OFFICER: BREDECK,JORDAN
ARRESTING OFFICER AGENCY: Lee Port Authority
LAST AGENCY INSPECTION: 03/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23016080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:34

SUBJECT'S LAST NAME: TOBIAS

DATE/TIME OF TEST: 03/18/2017 04:06:51

Diagnostics Check	OK	03:59
Air Blank	0.000	04:00
Control Test	0.079	04:00
Air Blank	0.000	04:01
Subject Sample # 1	0.111	04:02
Breath Volume	3.062	04:02
Air Blank	0.000	04:03
Air Blank	0.000	04:04
Subject Sample # 2	0.102	04:05
Breath Volume	1.546	04:05
Air Blank	0.000	04:05
Control Test	0.078	04:06
Air Blank	0.000	04:06
Diagnostics Check	OK	04:06

Data Download Date/Time: 04/12/2017 11:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALGER, MICHAEL, C
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: ALGER, MICHAEL
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 03/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23016080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:56

SUBJECT'S LAST NAME: ZUNIGA

DATE/TIME OF TEST: 03/18/2017 04:25:40

Diagnostics Check		OK	04:18
Air Blank		0.000	04:18
Control Test		0.078	04:19
Air Blank		0.000	04:19
Subject Sample	# 1	0.124	04:20
Breath Volume		1.785	04:20
Air Blank		0.000	04:21
Air Blank		0.000	04:23
Subject Sample	# 2	0.126	04:24
Breath Volume		2.902	04:24
Air Blank		0.000	04:24
Control Test		0.080	04:25
Air Blank		0.000	04:25
Diagnostics Check		OK	04:25

Data Download Date/Time: 04/12/2017 11:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NAYLOR, MAURICE, E
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: NAYLOR, MAURICE
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 03/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23016080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: RICE

DATE/TIME OF TEST: 03/19/2017 03:27:07

Diagnostics Check	OK	03:20
Air Blank	0.000	03:21
Control Test	0.081	03:21
Air Blank	0.000	03:22
Subject Sample # 1	0.060	03:22
Breath Volume	2.218	03:22
Air Blank	0.000	03:23
Air Blank	0.000	03:25
Subject Sample # 2	0.057	03:25
Breath Volume	1.738	03:25
Air Blank	0.000	03:26
Control Test	0.080	03:26
Air Blank	0.000	03:27
Diagnostics Check	OK	03:27

Data Download Date/Time: 04/12/2017 11:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH, JOSIAH, R
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: THABIT, S
ARRESTING OFFICER AGENCY: FL Gulf Coast UPD
LAST AGENCY INSPECTION: 03/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23016080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: FARRELL

DATE/TIME OF TEST: 03/19/2017 03:51:58

Diagnostics Check	OK	03:45
Air Blank	0.000	03:45
Control Test	0.080	03:46
Air Blank	0.000	03:46
Subject Sample # 1	0.097	03:47
Breath Volume	1.656	03:47
Air Blank	0.000	03:48
Air Blank	0.000	03:49
Subject Sample # 2	0.099	03:50
Breath Volume	2.121	03:50
Air Blank	0.000	03:51
Control Test	0.080	03:51
Air Blank	0.000	03:51
Diagnostics Check	OK	03:51

Data Download Date/Time: 04/12/2017 11:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALLOWAY,ERIC,B
BREATH TEST OPERATOR AGENCY: Lee Port Authority
ARRESTING OFFICER: BREDECK,JORDAN
ARRESTING OFFICER AGENCY: Lee Port Authority
LAST AGENCY INSPECTION: 03/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23016080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:39

SUBJECT'S LAST NAME: THEODORE

DATE/TIME OF TEST: 03/20/2017 03:10:10

Diagnostics Check	OK	03:03
Air Blank	0.000	03:03
Control Test	0.080	03:04
Air Blank	0.000	03:04
Subject Sample # 1	0.076	03:05
Breath Volume	2.578	03:05
Air Blank	0.000	03:06
Air Blank	0.000	03:07
Subject Sample # 2	0.072	03:08
Breath Volume	2.578	03:08
Air Blank	0.000	03:09
Control Test	0.079	03:09
Air Blank	0.000	03:10
Diagnostics Check	OK	03:10

Data Download Date/Time: 04/12/2017 11:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALLOWAY,ERIC,B
BREATH TEST OPERATOR AGENCY: Lee Port Authority
ARRESTING OFFICER: BREDECK,JORDAN
ARRESTING OFFICER AGENCY: Lee Port Authority
LAST AGENCY INSPECTION: 03/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23016080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 06:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:48

SUBJECT'S LAST NAME: TELLEZ

DATE/TIME OF TEST: 03/20/2017 07:15:26

Diagnostics Check	OK	07:09
Air Blank	0.000	07:09
Control Test	0.081	07:09
Air Blank	0.000	07:10
Subject Sample # 1	0.178	07:10
Breath Volume	1.593	07:10
Air Blank	0.000	07:11
Air Blank	0.000	07:13
Subject Sample # 2	0.173	07:13
Breath Volume	1.371	07:13
Air Blank	0.000	07:14
Control Test	0.079	07:14
Air Blank	0.000	07:15
Diagnostics Check	OK	07:15

Data Download Date/Time: 04/12/2017 11:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUNCAN,N,
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: DUNCAN,N
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 03/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23016080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:02

SUBJECT'S LAST NAME: JEWELL

DATE/TIME OF TEST: 03/21/2017 05:29:24

Diagnostics Check	OK	05:23
Air Blank	0.000	05:23
Control Test	0.080	05:24
Air Blank	0.000	05:24
Subject Sample # 1	0.000	05:25
Breath Volume	2.449	05:25
Air Blank	0.000	05:25
Air Blank	0.000	05:27
Subject Sample # 2	0.000	05:28
Breath Volume	2.441	05:28
Air Blank	0.000	05:28
Control Test	0.081	05:28
Air Blank	0.000	05:29
Diagnostics Check	OK	05:29

Data Download Date/Time: 04/12/2017 11:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUNCAN,N,
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: DOUGHERTY,M
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 03/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23016080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: FREITAS

DATE/TIME OF TEST: 03/25/2017 22:39:16

Diagnostics Check	OK	22:32
Air Blank	0.000	22:32
Control Test	0.080	22:33
Air Blank	0.000	22:33
Subject Sample # 1	0.162	22:34
Breath Volume	1.324	22:34
Air Blank	0.000	22:35
Air Blank	0.000	22:37
Subject Sample # 2	0.169	22:37
Breath Volume	1.734	22:37
Air Blank	0.000	22:38
Control Test	0.079	22:38
Air Blank	0.000	22:39
Diagnostics Check	OK	22:39

Data Download Date/Time: 04/12/2017 11:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUNCAN,N,
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: DUNCAN,N
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 03/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23016080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:08

SUBJECT'S LAST NAME: MORENO

DATE/TIME OF TEST: 03/25/2017 22:44:23

Diagnostics Check	OK	22:41	
Air Blank	0.000	22:41	
Control Test	0.079	22:41	
Air Blank	0.000	22:42	
Subject Sample	# 1	22:42	Interferent Detect
Breath Volume	1.921	22:42	
Air Blank	0.000	22:43	
Control Test	0.081	22:43	
Air Blank	0.000	22:44	
Diagnostics Check	OK	22:44	

Data Download Date/Time: 04/12/2017 11:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,ANDREW,R
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: SMITH,ANDREW
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 03/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23016080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:56

SUBJECT'S LAST NAME: MEAD

DATE/TIME OF TEST: 03/26/2017 03:21:46

Diagnostics Check	OK	03:18	
Air Blank	0.000	03:19	
Control Test	0.080	03:19	
Air Blank	0.000	03:20	
Subject Sample	# 1	03:20	Subject Test Refused
Breath Volume	0.000	03:20	
Air Blank	0.000	03:20	
Control Test	0.079	03:21	
Air Blank	0.000	03:21	
Diagnostics Check	OK	03:21	

Data Download Date/Time: 04/12/2017 11:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDIN, JEREMY, L
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: FINE, CHRISTOPHER
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 03/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23016080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 05:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:52

SUBJECT'S LAST NAME: AQUILERA

DATE/TIME OF TEST: 03/27/2017 06:22:34

Diagnostics Check		OK	06:16
Air Blank		0.000	06:16
Control Test		0.080	06:16
Air Blank		0.000	06:17
Subject Sample	# 1	0.108	06:18
Breath Volume		1.851	06:18
Air Blank		0.000	06:18
Air Blank		0.000	06:20
Subject Sample	# 2	0.108	06:20
Breath Volume		2.019	06:20
Air Blank		0.000	06:21
Control Test		0.078	06:21
Air Blank		0.000	06:22
Diagnostics Check		OK	06:22

Data Download Date/Time: 04/12/2017 11:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALGER, MICHAEL, C
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: ALGER, MICHAEL
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 03/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23016080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 05:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:46

SUBJECT'S LAST NAME: KINNEY

DATE/TIME OF TEST: 03/28/2017 06:22:23

Diagnostics Check	OK	06:15
Air Blank	0.000	06:16
Control Test	0.080	06:16
Air Blank	0.000	06:16
Subject Sample # 1	0.044	06:17
Breath Volume	2.117	06:17
Air Blank	0.000	06:18
Air Blank	0.000	06:19
Subject Sample # 2	0.046	06:20
Breath Volume	2.457	06:20
Air Blank	0.000	06:21
Control Test	0.081	06:21
Air Blank	0.000	06:22
Diagnostics Check	OK	06:22

Data Download Date/Time: 04/12/2017 11:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DECKER,STEPHEN,J
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: LUKASZ,C
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 03/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: SLOCK

DATE/TIME OF TEST: 03/04/2017 02:14:29

Diagnostics Check	OK	02:08
Air Blank	0.000	02:08
Control Test	0.079	02:09
Air Blank	0.000	02:09
Subject Sample # 1	0.140	02:09
Breath Volume	2.562	02:09
Air Blank	0.000	02:10
Air Blank	0.000	02:12
Subject Sample # 2	0.133	02:12
Breath Volume	2.167	02:12
Air Blank	0.000	02:13
Control Test	0.079	02:13
Air Blank	0.000	02:14
Diagnostics Check	OK	02:14

Data Download Date/Time: 04/04/2017 13:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NATSON,ANDRAE,M
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: GREEVES,JJ
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 03/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:01

SUBJECT'S LAST NAME: FRUMENTI

DATE/TIME OF TEST: 03/04/2017 03:35:46

Diagnostics Check	OK	03:25	
Air Blank	0.000	03:25	
Control Test	0.079	03:26	
Air Blank	0.000	03:26	
Subject Sample	# 1	03:29	No Sample Provided
Breath Volume	0.000	03:29	
Air Blank	0.000	03:30	
Air Blank	0.000	03:32	
Subject Sample	# 2	03:35	Sequence Aborted
Breath Volume	0.000	03:35	
Air Blank	0.000	03:35	

Data Download Date/Time: 04/04/2017 13:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WESTBERG,CHRISTOPHER,R
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: BAXTER,J
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 03/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 13:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:48

SUBJECT'S LAST NAME: SPARTA

DATE/TIME OF TEST: 03/04/2017 14:16:52

Diagnostics Check	OK	14:10
Air Blank	0.000	14:10
Control Test	0.079	14:11
Air Blank	0.000	14:11
Subject Sample # 1	0.125	14:12
Breath Volume	2.011	14:12
Air Blank	0.000	14:12
Air Blank	0.000	14:14
Subject Sample # 2	0.124	14:15
Breath Volume	2.964	14:15
Air Blank	0.000	14:15
Control Test	0.079	14:16
Air Blank	0.000	14:16
Diagnostics Check	OK	14:16

Data Download Date/Time: 04/04/2017 13:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VADNEY, MICHELE, M
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: MAHOLTE, S
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 03/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: ELLINGTON

DATE/TIME OF TEST: 03/05/2017 01:34:00

Diagnostics Check		OK	01:24
Air Blank		0.000	01:25
Control Test		0.079	01:25
Air Blank		0.000	01:25
Subject Sample	# 1	0.167	01:26
Breath Volume		1.367	01:26
Air Blank		0.000	01:27
Air Blank		0.000	01:29
Subject Sample	# 2	0.151	01:32
Breath Volume		1.632	01:32
Air Blank		0.000	01:33
Control Test		0.079	01:33
Air Blank		0.000	01:33
Diagnostics Check		OK	01:34

Data Download Date/Time: 04/04/2017 13:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DECKER,DAVID,P
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: ASHBY,N
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 03/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: YODER

DATE/TIME OF TEST: 03/06/2017 00:22:54

Diagnostics Check	OK	00:16
Air Blank	0.000	00:16
Control Test	0.079	00:17
Air Blank	0.000	00:17
Subject Sample # 1	0.139	00:18
Breath Volume	1.511	00:18
Air Blank	0.000	00:19
Air Blank	0.000	00:20
Subject Sample # 2	0.139	00:21
Breath Volume	1.371	00:21
Air Blank	0.000	00:21
Control Test	0.079	00:22
Air Blank	0.000	00:22
Diagnostics Check	OK	00:22

Data Download Date/Time: 04/04/2017 13:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KENNEDY,DANIEL,L
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: KENNEDY,DANIEL
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 03/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:26

SUBJECT'S LAST NAME: KLINGENSMITH

DATE/TIME OF TEST: 03/11/2017 00:58:29

Diagnostics Check	OK	00:51
Air Blank	0.000	00:52
Control Test	0.078	00:52
Air Blank	0.000	00:52
Subject Sample # 1	0.217	00:53
Breath Volume	2.171	00:53
Air Blank	0.000	00:54
Air Blank	0.000	00:56
Subject Sample # 2	0.202	00:56
Breath Volume	1.523	00:56
Air Blank	0.000	00:57
Control Test	0.078	00:57
Air Blank	0.000	00:58
Diagnostics Check	OK	00:58

Data Download Date/Time: 04/04/2017 13:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VONSOOSTEN,STEVEN,S
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: VONSOOSTEN,STEVEN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 03/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:28

SUBJECT'S LAST NAME: JENKINS

DATE/TIME OF TEST: 03/11/2017 03:59:03

Diagnostics Check		OK	03:52
Air Blank		0.000	03:53
Control Test		0.079	03:53
Air Blank		0.000	03:53
Subject Sample	# 1	0.122	03:54
Breath Volume		4.476	03:54
Air Blank		0.000	03:55
Air Blank		0.000	03:57
Subject Sample	# 2	0.120	03:57
Breath Volume		4.250	03:57
Air Blank		0.000	03:58
Control Test		0.078	03:58
Air Blank		0.000	03:58
Diagnostics Check		OK	03:59

Data Download Date/Time: 04/04/2017 13:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRATT,NICHOLAS,J
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: ROWLEY,C
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 03/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: ZAKRZEWSKI

DATE/TIME OF TEST: 03/11/2017 04:36:34

Diagnostics Check	OK	04:30
Air Blank	0.000	04:30
Control Test	0.079	04:30
Air Blank	0.000	04:31
Subject Sample # 1	0.146	04:32
Breath Volume	3.617	04:32
Air Blank	0.000	04:32
Air Blank	0.000	04:34
Subject Sample # 2	0.145	04:34
Breath Volume	3.132	04:34
Air Blank	0.000	04:35
Control Test	0.078	04:35
Air Blank	0.000	04:36
Diagnostics Check	OK	04:36

Data Download Date/Time: 04/04/2017 13:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIBEY,AMANDA,L
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: CORWIN,J
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 03/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:41

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 03/11/2017 05:10:40

Diagnostics Check	OK	05:03
Air Blank	0.000	05:04
Control Test	0.078	05:04
Air Blank	0.000	05:05
Subject Sample # 1	0.242	05:05
Breath Volume	1.289	05:05
Air Blank	0.000	05:06
Air Blank	0.000	05:08
Subject Sample # 2	0.235	05:09
Breath Volume	1.285	05:09
Air Blank	0.000	05:09
Control Test	0.078	05:10
Air Blank	0.000	05:10
Diagnostics Check	OK	05:10

Data Download Date/Time: 04/04/2017 13:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILIAN,MAYRA,G
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: MILLER,G
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 03/01/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:12

SUBJECT'S LAST NAME: MILIAN

DATE/TIME OF TEST: 03/11/2017 22:42:49

Diagnostics Check	OK	22:34
Air Blank	0.000	22:35
Control Test	0.079	22:35
Air Blank	0.000	22:35
Subject Sample # 1	0.238	22:36
Breath Volume	1.535	22:36
Air Blank	0.000	22:37
Air Blank	0.000	22:39
Subject Sample # 2	0.235	22:40
Breath Volume	1.750	22:40
Air Blank	0.000	22:41
Control Test	0.076	22:42
Air Blank	0.000	22:42
Diagnostics Check	OK	22:42

Data Download Date/Time: 04/04/2017 13:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MERTILUS,FREDELYNE,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: MILLER,G
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 03/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:12

SUBJECT'S LAST NAME: MILIAN

DATE/TIME OF TEST: 03/11/2017 22:50:41

Diagnostics Check	OK	22:44
Air Blank	0.000	22:44
Control Test	0.077	22:45
Air Blank	0.000	22:45
Subject Sample # 1	0.231	22:46
Breath Volume	1.371	22:46
Air Blank	0.000	22:46
Air Blank	0.000	22:48
Subject Sample # 2	0.234	22:49
Breath Volume	1.628	22:49
Air Blank	0.000	22:49
Control Test	0.076	22:50
Air Blank	0.000	22:50
Diagnostics Check	OK	22:50

Data Download Date/Time: 04/04/2017 13:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOPEZ,P,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: MERCADO,D
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 03/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: ALVAREZ SANCHEZ

DATE/TIME OF TEST: 03/12/2017 01:42:01

Diagnostics Check	OK	01:35
Air Blank	0.000	01:36
Control Test	0.078	01:36
Air Blank	0.000	01:37
Subject Sample # 1	0.193	01:37
Breath Volume	1.765	01:37
Air Blank	0.000	01:38
Air Blank	0.000	01:40
Subject Sample # 2	0.186	01:40
Breath Volume	2.488	01:40
Air Blank	0.000	01:41
Control Test	0.079	01:41
Air Blank	0.000	01:41
Diagnostics Check	OK	01:42

Data Download Date/Time: 04/04/2017 13:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROGERS, MATTHEW, L
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: DERRIG, J
ARRESTING OFFICER AGENCY: Marco Island PD
LAST AGENCY INSPECTION: 03/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: PARRISH

DATE/TIME OF TEST: 03/13/2017 00:07:49

Diagnostics Check		OK	00:01
Air Blank		0.000	00:01
Control Test		0.079	00:01
Air Blank		0.000	00:02
Subject Sample	# 1	0.258	00:02
Breath Volume		1.492	00:02
Air Blank		0.000	00:03
Air Blank		0.000	00:05
Subject Sample	# 2	0.271	00:06
Breath Volume		2.710	00:06
Air Blank		0.000	00:06
Control Test		0.079	00:07
Air Blank		0.000	00:07
Diagnostics Check		OK	00:07

Data Download Date/Time: 04/04/2017 13:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORNELIUS,TIM,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: CORNELIUS,TIM
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 03/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:42

SUBJECT'S LAST NAME: MACKETT

DATE/TIME OF TEST: 03/16/2017 02:10:43

Diagnostics Check	OK	02:04
Air Blank	0.000	02:04
Control Test	0.080	02:04
Air Blank	0.000	02:05
Subject Sample # 1	0.000	02:06
Breath Volume	1.343	02:06
Air Blank	0.000	02:06
Air Blank	0.000	02:08
Subject Sample # 2	0.000	02:09
Breath Volume	1.789	02:09
Air Blank	0.000	02:09
Control Test	0.079	02:10
Air Blank	0.000	02:10
Diagnostics Check	OK	02:10

Data Download Date/Time: 04/04/2017 13:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHRISTIANSEN, DAVID, M
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: CHRISTIANSEN, DAVID
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 03/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: FLAMBO

DATE/TIME OF TEST: 03/16/2017 03:45:55

Diagnostics Check	OK	03:34		
Air Blank	0.000	03:34		
Control Test	0.079	03:35		
Air Blank	0.000	03:35		
Subject Sample	# 1	03:38	No Sample Provided	
Breath Volume	0.000	03:38		
Air Blank	0.000	03:39		
Air Blank	0.000	03:41		
Subject Sample	# 2	0.074	03:44	Volume Not Met
Breath Volume	0.074	03:44		
Air Blank	0.000	03:45		
Control Test	0.079	03:45		
Air Blank	0.000	03:45		
Diagnostics Check	OK	03:45		

Data Download Date/Time: 04/04/2017 13:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHRISTIANSEN, DAVID, M
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: CHRISTIANSEN, DAVID
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 03/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: FLAMBO

DATE/TIME OF TEST: 03/16/2017 04:01:07

Diagnostics Check		OK	03:47	
Air Blank		0.000	03:47	
Control Test		0.079	03:48	
Air Blank		0.000	03:48	
Subject Sample	# 1	0.117	03:51	Volume Not Met
Breath Volume		0.156	03:51	
Air Blank		0.000	03:52	
Air Blank		0.000	03:54	
Subject Sample	# 2	0.138	03:55	
Breath Volume		1.726	03:55	
Air Blank		0.000	03:56	
Air Blank		0.000	03:57	
Subject Sample	# 3	0.149	03:59	
Breath Volume		1.707	03:59	
Air Blank		0.000	04:00	
Control Test		0.078	04:00	
Air Blank		0.000	04:01	
Diagnostics Check		OK	04:01	

Data Download Date/Time: 04/04/2017 13:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NOEL,YVES,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: WARGO,T
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 03/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: HONORAT

DATE/TIME OF TEST: 03/19/2017 04:23:13

Diagnostics Check	OK	04:16
Air Blank	0.000	04:16
Control Test	0.079	04:17
Air Blank	0.000	04:17
Subject Sample # 1	0.163	04:18
Breath Volume	2.410	04:18
Air Blank	0.000	04:18
Air Blank	0.000	04:20
Subject Sample # 2	0.173	04:21
Breath Volume	2.617	04:21
Air Blank	0.000	04:22
Control Test	0.080	04:22
Air Blank	0.000	04:23
Diagnostics Check	OK	04:23

Data Download Date/Time: 04/04/2017 13:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRAY, MICHAEL, A
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: ASHBY, NATALIE
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 03/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 05:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:25

SUBJECT'S LAST NAME: MANSFIELD

DATE/TIME OF TEST: 03/19/2017 06:56:07

Diagnostics Check	OK	06:49
Air Blank	0.000	06:50
Control Test	0.079	06:50
Air Blank	0.000	06:51
Subject Sample # 1	0.065	06:51
Breath Volume	1.140	06:51
Air Blank	0.000	06:52
Air Blank	0.000	06:54
Subject Sample # 2	0.061	06:54
Breath Volume	2.773	06:54
Air Blank	0.000	06:55
Control Test	0.079	06:55
Air Blank	0.000	06:56
Diagnostics Check	OK	06:56

Data Download Date/Time: 04/04/2017 13:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDUFFIE,B.,A
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: RODRIGUEZ,DAVID
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 03/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 15:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:50

SUBJECT'S LAST NAME: PEREZMORALES

DATE/TIME OF TEST: 03/19/2017 17:21:56

Diagnostics Check	OK	17:15
Air Blank	0.000	17:15
Control Test	0.079	17:15
Air Blank	0.000	17:16
Subject Sample # 1	0.244	17:17
Breath Volume	3.230	17:17
Air Blank	0.000	17:17
Air Blank	0.000	17:19
Subject Sample # 2	0.246	17:20
Breath Volume	2.406	17:20
Air Blank	0.000	17:21
Control Test	0.079	17:21
Air Blank	0.000	17:21
Diagnostics Check	OK	17:21

Data Download Date/Time: 04/04/2017 13:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ASHLINE,C,J
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: MCCOY,D
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 03/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: BUCKHAM

DATE/TIME OF TEST: 03/20/2017 01:15:01

Diagnostics Check		OK	01:08
Air Blank		0.000	01:08
Control Test		0.079	01:08
Air Blank		0.000	01:09
Subject Sample	# 1	0.275	01:10
Breath Volume		6.050	01:10
Air Blank		0.000	01:11
Air Blank		0.000	01:12
Subject Sample	# 2	0.269	01:13
Breath Volume		5.292	01:13
Air Blank		0.000	01:14
Control Test		0.078	01:14
Air Blank		0.000	01:14
Diagnostics Check		OK	01:15

Data Download Date/Time: 04/04/2017 13:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KENNEDY,DANIEL,L
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: KENNEDY,DANIEL
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 03/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 17:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:03

SUBJECT'S LAST NAME: MENSINGER

DATE/TIME OF TEST: 03/20/2017 18:34:00

Diagnostics Check	OK	18:27
Air Blank	0.000	18:27
Control Test	0.079	18:27
Air Blank	0.000	18:28
Subject Sample # 1	0.072	18:28
Breath Volume	1.429	18:28
Air Blank	0.000	18:29
Air Blank	0.000	18:31
Subject Sample # 2	0.073	18:32
Breath Volume	2.011	18:32
Air Blank	0.000	18:33
Control Test	0.079	18:33
Air Blank	0.000	18:33
Diagnostics Check	OK	18:34

Data Download Date/Time: 04/04/2017 13:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARRILLO,XAVIER,R
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: CORWIN,J
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 03/01/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:57

SUBJECT'S LAST NAME: BELLANE

DATE/TIME OF TEST: 03/20/2017 23:23:41

Diagnostics Check	OK	23:17
Air Blank	0.000	23:17
Control Test	0.079	23:18
Air Blank	0.000	23:18
Subject Sample # 1	0.000	23:19
Breath Volume	1.156	23:19
Air Blank	0.000	23:19
Air Blank	0.000	23:21
Subject Sample # 2	0.000	23:22
Breath Volume	1.148	23:22
Air Blank	0.000	23:22
Control Test	0.079	23:23
Air Blank	0.000	23:23
Diagnostics Check	OK	23:23

Data Download Date/Time: 04/04/2017 13:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUGO,JONATHAN,B
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: ASHBY,N
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 03/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 17:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:31

SUBJECT'S LAST NAME: SHEERAN

DATE/TIME OF TEST: 03/23/2017 19:02:13

Diagnostics Check	OK	18:55
Air Blank	0.000	18:56
Control Test	0.079	18:56
Air Blank	0.000	18:57
Subject Sample # 1	0.234	18:57
Breath Volume	2.484	18:57
Air Blank	0.000	18:58
Air Blank	0.000	19:00
Subject Sample # 2	0.241	19:00
Breath Volume	2.132	19:00
Air Blank	0.000	19:01
Control Test	0.079	19:01
Air Blank	0.000	19:02
Diagnostics Check	OK	19:02

Data Download Date/Time: 04/04/2017 13:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BASTON, MITCHELL, J
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: GREEVES, JAMES
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 03/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:46

SUBJECT'S LAST NAME: FURANNA

DATE/TIME OF TEST: 03/23/2017 22:16:59

Diagnostics Check	OK	22:10
Air Blank	0.000	22:11
Control Test	0.079	22:11
Air Blank	0.000	22:11
Subject Sample # 1	0.119	22:12
Breath Volume	2.605	22:12
Air Blank	0.000	22:13
Air Blank	0.000	22:14
Subject Sample # 2	0.118	22:15
Breath Volume	2.253	22:15
Air Blank	0.000	22:16
Control Test	0.079	22:16
Air Blank	0.000	22:16
Diagnostics Check	OK	22:16

Data Download Date/Time: 04/04/2017 13:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUTIERREZ,I,T
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: PICCOLI,M
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 03/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:48

SUBJECT'S LAST NAME: METCALF

DATE/TIME OF TEST: 03/24/2017 02:16:37

Diagnostics Check	OK	02:10
Air Blank	0.000	02:10
Control Test	0.079	02:11
Air Blank	0.000	02:11
Subject Sample # 1	0.248	02:12
Breath Volume	2.085	02:12
Air Blank	0.000	02:12
Air Blank	0.000	02:14
Subject Sample # 2	0.243	02:15
Breath Volume	1.582	02:15
Air Blank	0.000	02:15
Control Test	0.078	02:16
Air Blank	0.000	02:16
Diagnostics Check	OK	02:16

Data Download Date/Time: 04/04/2017 13:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CURTSINGER,CODY,A
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: WEEKENDER,W
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 03/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 18:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:35

SUBJECT'S LAST NAME: SIMEON

DATE/TIME OF TEST: 03/24/2017 19:10:40

Diagnostics Check	OK	19:04
Air Blank	0.000	19:05
Control Test	0.080	19:05
Air Blank	0.000	19:05
Subject Sample # 1	0.024	19:06
Breath Volume	2.539	19:06
Air Blank	0.000	19:06
Air Blank	0.000	19:08
Subject Sample # 2	0.025	19:09
Breath Volume	2.582	19:09
Air Blank	0.000	19:09
Control Test	0.080	19:10
Air Blank	0.000	19:10
Diagnostics Check	OK	19:10

Data Download Date/Time: 04/04/2017 13:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORNELIUS,TIM,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: CORNELUIS,TIM
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 03/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:39

SUBJECT'S LAST NAME: CASSESE

DATE/TIME OF TEST: 03/25/2017 04:11:31

Diagnostics Check	OK	04:04
Air Blank	0.000	04:05
Control Test	0.079	04:05
Air Blank	0.000	04:06
Subject Sample # 1	0.173	04:06
Breath Volume	2.500	04:06
Air Blank	0.000	04:07
Air Blank	0.000	04:09
Subject Sample # 2	0.167	04:09
Breath Volume	1.917	04:09
Air Blank	0.000	04:10
Control Test	0.080	04:10
Air Blank	0.000	04:11
Diagnostics Check	OK	04:11

Data Download Date/Time: 04/04/2017 13:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KENNEDY,DANIEL,L
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: KENNEDY,DANIEL
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 03/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: ANGELES

DATE/TIME OF TEST: 03/26/2017 04:25:33

Diagnostics Check	OK	04:18
Air Blank	0.000	04:19
Control Test	0.079	04:19
Air Blank	0.000	04:20
Subject Sample # 1	0.183	04:20
Breath Volume	2.835	04:20
Air Blank	0.000	04:21
Air Blank	0.000	04:23
Subject Sample # 2	0.171	04:24
Breath Volume	2.574	04:24
Air Blank	0.000	04:24
Control Test	0.078	04:24
Air Blank	0.000	04:25
Diagnostics Check	OK	04:25

Data Download Date/Time: 04/04/2017 13:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KENNEDY,DANIEL,L
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: KENNEDY,DANIEL
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 03/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:58

SUBJECT'S LAST NAME: CLARK

DATE/TIME OF TEST: 03/26/2017 22:28:20

Diagnostics Check	OK	22:21
Air Blank	0.000	22:22
Control Test	0.079	22:22
Air Blank	0.000	22:23
Subject Sample # 1	0.171	22:23
Breath Volume	1.855	22:23
Air Blank	0.000	22:24
Air Blank	0.000	22:26
Subject Sample # 2	0.177	22:26
Breath Volume	1.718	22:26
Air Blank	0.000	22:27
Control Test	0.079	22:27
Air Blank	0.000	22:28
Diagnostics Check	OK	22:28

Data Download Date/Time: 04/04/2017 13:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORNELIUS,TIM,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: CORNELIUS,TIM
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 03/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:24

SUBJECT'S LAST NAME: DILKES

DATE/TIME OF TEST: 03/27/2017 03:55:38

Diagnostics Check		OK	03:48
Air Blank		0.000	03:49
Control Test		0.079	03:49
Air Blank		0.000	03:50
Subject Sample	# 1	0.254	03:50
Breath Volume		3.429	03:50
Air Blank		0.000	03:51
Air Blank		0.000	03:53
Subject Sample	# 2	0.261	03:54
Breath Volume		3.300	03:54
Air Blank		0.000	03:54
Control Test		0.078	03:55
Air Blank		0.000	03:55
Diagnostics Check		OK	03:55

Data Download Date/Time: 04/04/2017 13:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUONPANE,VINCENT,D
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: MILLER,K
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 03/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:49

SUBJECT'S LAST NAME: VASQUEZ

DATE/TIME OF TEST: 03/27/2017 04:21:34

Diagnostics Check	OK	04:14
Air Blank	0.000	04:14
Control Test	0.079	04:14
Air Blank	0.000	04:15
Subject Sample # 1	0.096	04:16
Breath Volume	1.742	04:16
Air Blank	0.000	04:17
Air Blank	0.000	04:18
Subject Sample # 2	0.094	04:20
Breath Volume	1.492	04:20
Air Blank	0.000	04:20
Control Test	0.079	04:20
Air Blank	0.000	04:21
Diagnostics Check	OK	04:21

Data Download Date/Time: 04/04/2017 13:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARPENTER,CHRISTOPHER,J
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: CARPENTER,CHRISTOPHER
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 03/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: PROB/PAROL
ARREST TIME: 19:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:22

SUBJECT'S LAST NAME: IRWIN

DATE/TIME OF TEST: 03/27/2017 20:52:30

Diagnostics Check	OK	20:46
Air Blank	0.000	20:46
Control Test	0.079	20:46
Air Blank	0.000	20:47
Subject Sample # 1	0.180	20:48
Breath Volume	1.601	20:48
Air Blank	0.000	20:48
Air Blank	0.000	20:50
Subject Sample # 2	0.173	20:50
Breath Volume	1.601	20:50
Air Blank	0.000	20:51
Control Test	0.078	20:51
Air Blank	0.000	20:52
Diagnostics Check	OK	20:52

Data Download Date/Time: 04/04/2017 13:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARPENTER,CHRISTOPHER,J
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: DYER,BILL
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 03/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:38

SUBJECT'S LAST NAME: BOSLEY

DATE/TIME OF TEST: 03/27/2017 22:06:06

Diagnostics Check	OK	21:59
Air Blank	0.000	22:00
Control Test	0.079	22:00
Air Blank	0.000	22:01
Subject Sample # 1	0.192	22:01
Breath Volume	2.156	22:01
Air Blank	0.000	22:02
Air Blank	0.000	22:04
Subject Sample # 2	0.185	22:04
Breath Volume	1.816	22:04
Air Blank	0.000	22:05
Control Test	0.079	22:05
Air Blank	0.000	22:06
Diagnostics Check	OK	22:06

Data Download Date/Time: 04/04/2017 13:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GEORGE,SHANNON,N
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: NOVA,M
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 03/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: GARREN

DATE/TIME OF TEST: 03/29/2017 03:37:48

Diagnostics Check		OK	03:31
Air Blank		0.000	03:31
Control Test		0.079	03:31
Air Blank		0.000	03:32
Subject Sample	# 1	0.203	03:32
Breath Volume		2.027	03:32
Air Blank		0.000	03:33
Air Blank		0.000	03:35
Subject Sample	# 2	0.208	03:36
Breath Volume		2.132	03:36
Air Blank		0.000	03:36
Control Test		0.079	03:37
Air Blank		0.000	03:37
Diagnostics Check		OK	03:37

Data Download Date/Time: 04/04/2017 13:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORNELIUS,TIM,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: CORNELIUS,TIM
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 03/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:02

SUBJECT'S LAST NAME: STULL

DATE/TIME OF TEST: 03/30/2017 01:33:59

Diagnostics Check	OK	01:27
Air Blank	0.000	01:27
Control Test	0.080	01:27
Air Blank	0.000	01:28
Subject Sample # 1	0.180	01:29
Breath Volume	2.343	01:29
Air Blank	0.000	01:29
Air Blank	0.000	01:31
Subject Sample # 2	0.181	01:32
Breath Volume	2.238	01:32
Air Blank	0.000	01:33
Control Test	0.079	01:33
Air Blank	0.000	01:33
Diagnostics Check	OK	01:33

Data Download Date/Time: 04/04/2017 13:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAN-KITA,JACQUES,J
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: PAN-KITA,JACQUES
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 03/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 13:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:56

SUBJECT'S LAST NAME: BELKNAP

DATE/TIME OF TEST: 03/30/2017 14:40:48

Diagnostics Check		OK	14:29	
Air Blank		0.000	14:29	
Control Test		0.080	14:29	
Air Blank		0.000	14:30	
Subject Sample	# 1	0.210	14:33	Volume Not Met
Breath Volume		0.183	14:33	
Air Blank		0.000	14:34	
Air Blank		0.000	14:35	
Subject Sample	# 2	0.176	14:39	Volume Not Met
Breath Volume		0.335	14:39	
Air Blank		0.000	14:39	
Control Test		0.078	14:40	
Air Blank		0.000	14:40	
Diagnostics Check		OK	14:40	

Data Download Date/Time: 04/04/2017 13:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAN-KITA,JACQUES,J
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: PAN-KITA,JACQUES
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 03/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 13:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:56

SUBJECT'S LAST NAME: BELKNAP

DATE/TIME OF TEST: 03/30/2017 14:58:27

Diagnostics Check		OK	14:43	
Air Blank		0.000	14:43	
Control Test		0.078	14:44	
Air Blank		0.000	14:44	
Subject Sample	# 1	0.185	14:47	Volume Not Met
Breath Volume		0.640	14:47	
Air Blank		0.000	14:48	
Air Blank		0.000	14:50	
Subject Sample	# 2	0.216	14:53	
Breath Volume		1.648	14:53	
Air Blank		0.000	14:53	
Air Blank		0.000	14:55	
Subject Sample	# 3	0.209	14:56	
Breath Volume		1.257	14:56	
Air Blank		0.000	14:57	
Control Test		0.077	14:57	
Air Blank		0.000	14:58	
Diagnostics Check		OK	14:58	

Data Download Date/Time: 04/04/2017 13:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DRISCOLL,JAMES,P
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: DRISCOLL,JAMES
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 03/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: BECHLIVANIS-GRANJA

DATE/TIME OF TEST: 03/31/2017 01:59:00

Diagnostics Check		OK	01:52
Air Blank		0.000	01:52
Control Test		0.079	01:52
Air Blank		0.000	01:53
Subject Sample	# 1	0.227	01:54
Breath Volume		2.218	01:54
Air Blank		0.000	01:55
Air Blank		0.000	01:56
Subject Sample	# 2	0.218	01:57
Breath Volume		3.058	01:57
Air Blank		0.000	01:58
Control Test		0.079	01:58
Air Blank		0.000	01:58
Diagnostics Check		OK	01:59

Data Download Date/Time: 04/04/2017 13:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000946

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMERON,CELESTE,M
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: EPARDS,R
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 03/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 08:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:47

SUBJECT'S LAST NAME: JANISZEWSKA

DATE/TIME OF TEST: 03/04/2017 11:13:24

Diagnostics Check	OK	11:07
Air Blank	0.000	11:07
Control Test	0.079	11:08
Air Blank	0.000	11:08
Subject Sample # 1	0.000	11:09
Breath Volume	3.539	11:09
Air Blank	0.000	11:09
Air Blank	0.000	11:11
Subject Sample # 2	0.000	11:12
Breath Volume	3.019	11:12
Air Blank	0.000	11:12
Control Test	0.079	11:12
Air Blank	0.000	11:13
Diagnostics Check	OK	11:13

Data Download Date/Time: 04/19/2017 10:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000946

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCLAIN,CHRISTOPHER,M
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: SCHULTE,A
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 03/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:07

SUBJECT'S LAST NAME: ROCK

DATE/TIME OF TEST: 03/05/2017 02:42:29

Diagnostics Check		OK	02:28	
Air Blank		0.000	02:28	
Control Test		0.079	02:28	
Air Blank		0.000	02:29	
Subject Sample	# 1	0.168	02:32	Volume Not Met
Breath Volume		0.035	02:32	
Air Blank		0.000	02:33	
Air Blank		0.000	02:34	
Subject Sample	# 2	0.148	02:36	
Breath Volume		2.402	02:36	
Air Blank		0.000	02:37	
Air Blank		0.000	02:39	
Subject Sample	# 3	0.161	02:41	
Breath Volume		1.550	02:41	
Air Blank		0.000	02:41	
Control Test		0.078	02:41	
Air Blank		0.000	02:42	
Diagnostics Check		OK	02:42	

Data Download Date/Time: 04/19/2017 10:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000946

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCLAIN,CHRISTOPHER,M
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: DOOLEY,B
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 03/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: BAILEY

DATE/TIME OF TEST: 03/05/2017 03:47:01

Diagnostics Check		OK	03:39
Air Blank		0.000	03:39
Control Test		0.078	03:39
Air Blank		0.000	03:40
Subject Sample	# 1	0.208	03:41
Breath Volume		2.472	03:41
Air Blank		0.000	03:42
Air Blank		0.000	03:44
Subject Sample	# 2	0.202	03:45
Breath Volume		1.164	03:45
Air Blank		0.000	03:46
Control Test		0.079	03:46
Air Blank		0.000	03:46
Diagnostics Check		OK	03:47

Data Download Date/Time: 04/19/2017 10:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000946

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHAPMAN,VANESSA,N
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: DOOLEY,BROOKE
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 03/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:41

SUBJECT'S LAST NAME: WALSH

DATE/TIME OF TEST: 03/08/2017 22:11:40

Diagnostics Check	OK	22:04
Air Blank	0.000	22:04
Control Test	0.079	22:04
Air Blank	0.000	22:05
Subject Sample # 1	0.354	22:06
Breath Volume	2.527	22:06
Air Blank	0.000	22:07
Air Blank	0.000	22:09
Subject Sample # 2	0.342	22:10
Breath Volume	2.042	22:10
Air Blank	0.000	22:10
Control Test	0.078	22:11
Air Blank	0.000	22:11
Diagnostics Check	OK	22:11

Data Download Date/Time: 04/19/2017 10:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000946

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FISHER,SCOTT,R
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: SELLA,STEVEN
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 03/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: HORNBACK

DATE/TIME OF TEST: 03/11/2017 02:25:45

Diagnostics Check	OK	02:16
Air Blank	0.000	02:16
Control Test	0.079	02:17
Air Blank	0.000	02:17
Subject Sample # 1	0.086	02:20
Breath Volume	2.421	02:20
Air Blank	0.000	02:21
Air Blank	0.000	02:22
Subject Sample # 2	0.086	02:24
Breath Volume	2.894	02:24
Air Blank	0.000	02:24
Control Test	0.079	02:25
Air Blank	0.000	02:25
Diagnostics Check	OK	02:25

Data Download Date/Time: 04/19/2017 10:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000946

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHEPARD,CHAD,A
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: HARTMANSTOFER,CHRIS
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 03/01/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 15:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:25

SUBJECT'S LAST NAME: SWORE

DATE/TIME OF TEST: 03/13/2017 16:53:40

Diagnostics Check	OK	16:47
Air Blank	0.000	16:47
Control Test	0.080	16:48
Air Blank	0.000	16:48
Subject Sample # 1	0.150	16:49
Breath Volume	1.777	16:49
Air Blank	0.000	16:49
Air Blank	0.000	16:51
Subject Sample # 2	0.154	16:52
Breath Volume	1.785	16:52
Air Blank	0.000	16:52
Control Test	0.079	16:53
Air Blank	0.000	16:53
Diagnostics Check	OK	16:53

Data Download Date/Time: 04/19/2017 10:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000946

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHAPMAN,VANESSA,N
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: LEUTHNER,NICK
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 03/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:08

SUBJECT'S LAST NAME: COWLES

DATE/TIME OF TEST: 03/15/2017 01:46:07

Diagnostics Check		OK	01:31	
Air Blank		0.000	01:31	
Control Test		0.079	01:32	
Air Blank		0.000	01:32	
Subject Sample	# 1	0.270	01:34	
Breath Volume		1.320	01:34	
Air Blank		0.000	01:34	
Air Blank		0.000	01:36	
Subject Sample	# 2	0.254	01:39	Volume Not Met
Breath Volume		0.812	01:39	
Air Blank		0.000	01:40	
Air Blank		0.000	01:42	
Subject Sample	# 3	0.268	01:44	
Breath Volume		1.550	01:44	
Air Blank		0.000	01:45	
Control Test		0.077	01:45	
Air Blank		0.000	01:46	
Diagnostics Check		OK	01:46	

Data Download Date/Time: 04/19/2017 10:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000946

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHAPMAN,VANESSA,N
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: STAUB,DAN
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 03/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:16

SUBJECT'S LAST NAME: BRUCK

DATE/TIME OF TEST: 03/15/2017 03:44:47

Diagnostics Check	OK	03:36
Air Blank	0.000	03:37
Control Test	0.079	03:37
Air Blank	0.000	03:38
Subject Sample # 1	0.102	03:40
Breath Volume	1.785	03:40
Air Blank	0.000	03:40
Air Blank	0.000	03:42
Subject Sample # 2	0.101	03:43
Breath Volume	1.894	03:43
Air Blank	0.000	03:43
Control Test	0.079	03:44
Air Blank	0.000	03:44
Diagnostics Check	OK	03:44

Data Download Date/Time: 04/19/2017 10:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000946

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PATTON,JOHN,G
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: LEUTHNER,NICHOLAS
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 03/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:38

SUBJECT'S LAST NAME: PEASE

DATE/TIME OF TEST: 03/16/2017 04:06:17

Diagnostics Check	OK	03:58
Air Blank	0.000	03:59
Control Test	0.080	03:59
Air Blank	0.000	03:59
Subject Sample # 1	0.142	04:01
Breath Volume	2.234	04:01
Air Blank	0.000	04:02
Air Blank	0.000	04:04
Subject Sample # 2	0.153	04:04
Breath Volume	2.664	04:04
Air Blank	0.000	04:05
Control Test	0.079	04:05
Air Blank	0.000	04:06
Diagnostics Check	OK	04:06

Data Download Date/Time: 04/19/2017 10:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000946

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCLAIN,CHRISTOPHER,M
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: QUEGAN,T
ARRESTING OFFICER AGENCY: Punta Gorda PD
LAST AGENCY INSPECTION: 03/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:59

SUBJECT'S LAST NAME: PARROTT

DATE/TIME OF TEST: 03/17/2017 23:26:16

Diagnostics Check	OK	23:20
Air Blank	0.000	23:20
Control Test	0.079	23:20
Air Blank	0.000	23:21
Subject Sample # 1	0.092	23:21
Breath Volume	3.785	23:21
Air Blank	0.000	23:22
Air Blank	0.000	23:24
Subject Sample # 2	0.092	23:24
Breath Volume	3.070	23:24
Air Blank	0.000	23:25
Control Test	0.079	23:25
Air Blank	0.000	23:26
Diagnostics Check	OK	23:26

Data Download Date/Time: 04/19/2017 10:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000946

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHAPMAN,VANESSA,N
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: SCHULTE,ANDREW
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 03/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:33

SUBJECT'S LAST NAME: COUTURE

DATE/TIME OF TEST: 03/18/2017 01:01:20

Diagnostics Check	OK	00:54
Air Blank	0.000	00:55
Control Test	0.079	00:55
Air Blank	0.000	00:55
Subject Sample # 1	0.168	00:56
Breath Volume	1.218	00:56
Air Blank	0.000	00:57
Air Blank	0.000	00:59
Subject Sample # 2	0.171	00:59
Breath Volume	1.593	00:59
Air Blank	0.000	01:00
Control Test	0.079	01:00
Air Blank	0.000	01:01
Diagnostics Check	OK	01:01

Data Download Date/Time: 04/19/2017 10:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000946

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCLAIN,CHRISTOPHER,M
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: FREITAS,A
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 03/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:34

SUBJECT'S LAST NAME: PANNHORST

DATE/TIME OF TEST: 03/18/2017 02:07:21

Diagnostics Check	OK	02:00
Air Blank	0.000	02:00
Control Test	0.079	02:00
Air Blank	0.000	02:01
Subject Sample # 1	0.057	02:02
Breath Volume	2.074	02:02
Air Blank	0.000	02:03
Air Blank	0.000	02:04
Subject Sample # 2	0.056	02:05
Breath Volume	1.421	02:05
Air Blank	0.000	02:06
Control Test	0.079	02:06
Air Blank	0.000	02:07
Diagnostics Check	OK	02:07

Data Download Date/Time: 04/19/2017 10:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000946

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHAPMAN,VANESSA,N
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: QUEGAN,T
ARRESTING OFFICER AGENCY: Punta Gorda PD
LAST AGENCY INSPECTION: 03/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:56

SUBJECT'S LAST NAME: PAULK

DATE/TIME OF TEST: 03/18/2017 04:25:53

Diagnostics Check	OK	04:19
Air Blank	0.000	04:20
Control Test	0.080	04:20
Air Blank	0.000	04:20
Subject Sample # 1	0.139	04:21
Breath Volume	1.597	04:21
Air Blank	0.000	04:21
Air Blank	0.000	04:23
Subject Sample # 2	0.147	04:24
Breath Volume	1.507	04:24
Air Blank	0.000	04:25
Control Test	0.079	04:25
Air Blank	0.000	04:25
Diagnostics Check	OK	04:25

Data Download Date/Time: 04/19/2017 10:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000946

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KLOS,LINDSAY,M
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: TRUFAN,OFFICER
ARRESTING OFFICER AGENCY: Punta Gorda PD
LAST AGENCY INSPECTION: 03/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: FRAME

DATE/TIME OF TEST: 03/19/2017 03:01:21

Diagnostics Check	OK	02:46	
Air Blank	0.000	02:47	
Control Test	0.080	02:47	
Air Blank	0.000	02:47	
Subject Sample	# 1 0.165	02:50	
Breath Volume	1.109	02:50	
Air Blank	0.000	02:50	
Air Blank	0.000	02:52	
Subject Sample	# 2 0.165	02:55	Volume Not Met
Breath Volume	0.476	02:55	
Air Blank	0.000	02:56	
Air Blank	0.000	02:58	
Subject Sample	# 3 0.169	02:59	
Breath Volume	2.269	02:59	
Air Blank	0.000	03:00	
Control Test	0.078	03:00	
Air Blank	0.000	03:01	
Diagnostics Check	OK	03:01	

Data Download Date/Time: 04/19/2017 10:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000946

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KLOS,LINDSAY,M
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: GRIFFIN,DEPUTY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 03/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:47

SUBJECT'S LAST NAME: GENRTY II

DATE/TIME OF TEST: 03/21/2017 00:16:21

Diagnostics Check	OK	00:08
Air Blank	0.000	00:08
Control Test	0.080	00:09
Air Blank	0.000	00:09
Subject Sample # 1	0.041	00:11
Breath Volume	2.296	00:11
Air Blank	0.000	00:11
Air Blank	0.000	00:13
Subject Sample # 2	0.041	00:14
Breath Volume	2.156	00:14
Air Blank	0.000	00:15
Control Test	0.080	00:15
Air Blank	0.000	00:16
Diagnostics Check	OK	00:16

Data Download Date/Time: 04/19/2017 10:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000946

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FISHE,SCOTT,R
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: ROSE,DEPUTY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 03/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: GROW

DATE/TIME OF TEST: 03/26/2017 03:21:44

Diagnostics Check	OK	03:15
Air Blank	0.000	03:15
Control Test	0.080	03:15
Air Blank	0.000	03:16
Subject Sample # 1	0.206	03:17
Breath Volume	3.152	03:17
Air Blank	0.000	03:17
Air Blank	0.000	03:19
Subject Sample # 2	0.203	03:20
Breath Volume	3.164	03:20
Air Blank	0.000	03:20
Control Test	0.079	03:21
Air Blank	0.000	03:21
Diagnostics Check	OK	03:21

Data Download Date/Time: 04/19/2017 10:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000946

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMYRNIOS,MANUEL,J
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: SMYRNIOS,MANUEL
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 03/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: ROGATE

DATE/TIME OF TEST: 03/28/2017 04:47:02

Diagnostics Check	OK	04:41
Air Blank	0.000	04:41
Control Test	0.079	04:41
Air Blank	0.000	04:42
Subject Sample # 1	0.034	04:42
Breath Volume	1.371	04:42
Air Blank	0.000	04:43
Air Blank	0.000	04:45
Subject Sample # 2	0.038	04:45
Breath Volume	2.191	04:45
Air Blank	0.000	04:46
Control Test	0.079	04:46
Air Blank	0.000	04:46
Diagnostics Check	OK	04:47

Data Download Date/Time: 04/19/2017 10:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000946

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPBELL,J,M
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: RESTY,J
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 03/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 06:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:20

SUBJECT'S LAST NAME: JACOBSON

DATE/TIME OF TEST: 03/30/2017 07:03:17

Diagnostics Check		OK	06:50	
Air Blank		0.000	06:51	
Control Test		0.079	06:51	
Air Blank		0.000	06:51	
Subject Sample	# 1	0.071	06:55	Volume Not Met
Breath Volume		0.968	06:55	
Air Blank		0.000	06:55	
Air Blank		0.000	06:57	
Subject Sample	# 2	0.075	06:58	
Breath Volume		1.117	06:58	
Air Blank		0.000	06:59	
Air Blank		0.000	07:01	
Subject Sample	# 3	0.072	07:01	
Breath Volume		1.375	07:01	
Air Blank		0.000	07:02	
Control Test		0.077	07:02	
Air Blank		0.000	07:03	
Diagnostics Check		OK	07:03	

Data Download Date/Time: 04/19/2017 10:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000947

INSTRUMENT REGISTERED TO: GLADES COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HICKS,CHRISTINA,L
BREATH TEST OPERATOR AGENCY: Glades County SO
ARRESTING OFFICER: CLENDENING,JAMES
ARRESTING OFFICER AGENCY: Glades County SO
LAST AGENCY INSPECTION: 02/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04216080A3
DRY GAS EXPIRY DATE: 03/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 03/04/2017 01:46:18

Diagnostics Check	OK	01:38
Air Blank	0.000	01:39
Control Test	0.078	01:39
Air Blank	0.000	01:40
Subject Sample # 1	0.142	01:41
Breath Volume	1.683	01:41
Air Blank	0.000	01:42
Air Blank	0.000	01:43
Subject Sample # 2	0.145	01:44
Breath Volume	2.246	01:44
Air Blank	0.000	01:45
Control Test	0.076	01:45
Air Blank	0.000	01:46
Diagnostics Check	OK	01:46

Data Download Date/Time: 03/31/2017 10:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000947

INSTRUMENT REGISTERED TO: GLADES COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TOMS,DONNA,J
BREATH TEST OPERATOR AGENCY: Glades County SO
ARRESTING OFFICER: CLENDENING,JAMES
ARRESTING OFFICER AGENCY: Glades County SO
LAST AGENCY INSPECTION: 02/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04216080A3
DRY GAS EXPIRY DATE: 03/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:16

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 03/05/2017 05:03:36

Diagnostics Check		OK	04:49	
Air Blank		0.000	04:49	
Control Test		0.078	04:50	
Air Blank		0.000	04:50	
Subject Sample	# 1	0.107	04:53	Volume Not Met
Breath Volume		1.000	04:53	
Air Blank		0.000	04:54	
Air Blank		0.000	04:56	
Subject Sample	# 2	0.106	04:57	
Breath Volume		1.441	04:57	
Air Blank		0.000	04:57	
Air Blank		0.000	04:59	
Subject Sample	# 3	0.097	05:01	
Breath Volume		1.542	05:01	
Air Blank		0.000	05:02	
Control Test		0.077	05:02	
Air Blank		0.000	05:03	
Diagnostics Check		OK	05:03	

Data Download Date/Time: 03/31/2017 10:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000947

INSTRUMENT REGISTERED TO: GLADES COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLENDENING,JAMES,T
BREATH TEST OPERATOR AGENCY: Glades County SO
ARRESTING OFFICER: CLENDENING,JAMES
ARRESTING OFFICER AGENCY: Glades County SO
LAST AGENCY INSPECTION: 02/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04216080A3
DRY GAS EXPIRY DATE: 03/05/2018
VIOLATION CODE: ADMIN
ARREST TIME: 00:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: HERRERA

DATE/TIME OF TEST: 03/08/2017 01:40:09

Diagnostics Check	OK	01:31
Air Blank	0.000	01:31
Control Test	0.078	01:31
Air Blank	0.000	01:32
Subject Sample # 1	0.098	01:34
Breath Volume	1.535	01:34
Air Blank	0.000	01:35
Air Blank	0.000	01:36
Subject Sample # 2	0.097	01:38
Breath Volume	1.382	01:38
Air Blank	0.000	01:39
Control Test	0.077	01:39
Air Blank	0.000	01:40
Diagnostics Check	OK	01:40

Data Download Date/Time: 03/31/2017 10:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000947

INSTRUMENT REGISTERED TO: GLADES COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLENDENING,JAMES,T
BREATH TEST OPERATOR AGENCY: Glades County SO
ARRESTING OFFICER: CLENDENING,JAMES
ARRESTING OFFICER AGENCY: Glades County SO
LAST AGENCY INSPECTION: 02/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04216080A3
DRY GAS EXPIRY DATE: 03/05/2018
VIOLATION CODE: ADMIN
ARREST TIME: 00:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: HERRERA

DATE/TIME OF TEST: 03/08/2017 03:53:28

Diagnostics Check	OK	03:46
Air Blank	0.000	03:47
Control Test	0.078	03:47
Air Blank	0.000	03:47
Subject Sample # 1	0.051	03:48
Breath Volume	1.968	03:48
Air Blank	0.000	03:49
Air Blank	0.000	03:51
Subject Sample # 2	0.046	03:51
Breath Volume	1.621	03:51
Air Blank	0.000	03:52
Control Test	0.075	03:52
Air Blank	0.000	03:53
Diagnostics Check	OK	03:53

Data Download Date/Time: 03/31/2017 10:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000947

INSTRUMENT REGISTERED TO: GLADES COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HICKS,CHRISTINA,L
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: WARENSFORD,AUSTIN
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 02/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04216080A3
DRY GAS EXPIRY DATE: 03/05/2018
VIOLATION CODE: DUI
ARREST TIME: 12:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:43

SUBJECT'S LAST NAME: MENDOZA

DATE/TIME OF TEST: 03/16/2017 15:14:52

Diagnostics Check	OK	15:09	
Air Blank	0.000	15:09	
Control Test	0.078	15:09	
Air Blank	0.000	15:10	
Subject Sample	# 1	15:12	Subject Test Refused
Breath Volume	0.250	15:12	
Air Blank	0.000	15:13	
Control Test	0.077	15:14	
Air Blank	0.000	15:14	
Diagnostics Check	OK	15:14	

Data Download Date/Time: 03/31/2017 10:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000949

INSTRUMENT REGISTERED TO: TALLAHASSEE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHULZE,JON,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: SCHULZE,JON
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 02/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:13

SUBJECT'S LAST NAME: VON STEIN

DATE/TIME OF TEST: 03/01/2017 00:50:14

Diagnostics Check		OK	00:34	
Air Blank		0.000	00:34	
Control Test		0.078	00:34	
Air Blank		0.000	00:35	
Subject Sample	# 1	0.140	00:38	Volume Not Met
Breath Volume		0.078	00:38	
Air Blank		0.000	00:39	
Air Blank		0.000	00:40	
Subject Sample	# 2	0.144	00:43	
Breath Volume		1.449	00:43	
Air Blank		0.000	00:44	
Air Blank		0.000	00:45	
Subject Sample	# 3	0.137	00:48	
Breath Volume		1.882	00:48	
Air Blank		0.000	00:49	
Control Test		0.079	00:49	
Air Blank		0.000	00:50	
Diagnostics Check		OK	00:50	

Data Download Date/Time: 03/24/2017 15:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000949

INSTRUMENT REGISTERED TO: TALLAHASSEE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROY,G,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: ROY,G
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 03/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416090A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:28

SUBJECT'S LAST NAME: BAKER JR

DATE/TIME OF TEST: 03/26/2017 00:05:46

Diagnostics Check	OK	23:58
Air Blank	0.000	23:58
Control Test	0.083	23:59
Air Blank	0.000	23:59
Subject Sample # 1	0.140	00:00
Breath Volume	1.128	00:00
Air Blank	0.000	00:01
Air Blank	0.000	00:03
Subject Sample # 2	0.156	00:04
Breath Volume	2.082	00:04
Air Blank	0.000	00:04
Control Test	0.083	00:05
Air Blank	0.000	00:05
Diagnostics Check	OK	00:05

Data Download Date/Time: 04/28/2017 10:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000952

INSTRUMENT REGISTERED TO: FRANKLIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAM,RICHARD,A
BREATH TEST OPERATOR AGENCY: Franklin County SO
ARRESTING OFFICER: HAM,RICHARD
ARRESTING OFFICER AGENCY: Franklin County SO
LAST AGENCY INSPECTION: 02/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1841081
DRY GAS EXPIRY DATE: 04/08/2018
VIOLATION CODE: DUI
ARREST TIME: 19:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:17

SUBJECT'S LAST NAME: AARON

DATE/TIME OF TEST: 03/19/2017 20:47:22

Diagnostics Check	OK	20:41
Air Blank	0.000	20:41
Control Test	0.082	20:41
Air Blank	0.000	20:42
Subject Sample # 1	0.092	20:42
Breath Volume	4.574	20:42
Air Blank	0.000	20:43
Air Blank	0.000	20:45
Subject Sample # 2	0.085	20:45
Breath Volume	4.136	20:45
Air Blank	0.000	20:46
Control Test	0.080	20:46
Air Blank	0.000	20:47
Diagnostics Check	OK	20:47

Data Download Date/Time: 03/22/2017 15:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000952

INSTRUMENT REGISTERED TO: FRANKLIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBB,THOMAS,J
BREATH TEST OPERATOR AGENCY: Franklin County SO
ARRESTING OFFICER: MARTINA,GARY
ARRESTING OFFICER AGENCY: Franklin County SO
LAST AGENCY INSPECTION: 03/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1841081
DRY GAS EXPIRY DATE: 04/08/2018
VIOLATION CODE: DUI
ARREST TIME: 12:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:15

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 03/28/2017 13:04:28

Diagnostics Check	OK	13:01	
Air Blank	0.000	13:02	
Control Test	0.080	13:02	
Air Blank	0.000	13:02	
Subject Sample	# 1	13:03	Subject Test Refused
Breath Volume	0.000	13:03	
Air Blank	0.000	13:03	
Control Test	0.080	13:03	
Air Blank	0.000	13:04	
Diagnostics Check	OK	13:04	

Data Download Date/Time: 04/20/2017 09:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000955

INSTRUMENT REGISTERED TO: CLEWISTON P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHAMBERLAIN,STEVEN,M
BREATH TEST OPERATOR AGENCY: Clewiston PD
ARRESTING OFFICER: HODGE,KYLER
ARRESTING OFFICER AGENCY: Clewiston PD
LAST AGENCY INSPECTION: 03/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:49

SUBJECT'S LAST NAME: BORMUTH

DATE/TIME OF TEST: 03/03/2017 03:25:52

Diagnostics Check		OK	03:10	
Air Blank		0.000	03:11	
Control Test		0.081	03:11	
Air Blank		0.000	03:11	
Subject Sample	# 1	0.126	03:15	Volume Not Met
Breath Volume		0.593	03:15	
Air Blank		0.000	03:15	
Air Blank		0.000	03:17	
Subject Sample	# 2	0.118	03:18	
Breath Volume		1.808	03:18	
Air Blank		0.000	03:19	
Air Blank		0.000	03:21	
Subject Sample	# 3	0.098	03:24	Volume Not Met
Breath Volume		0.328	03:24	
Air Blank		0.000	03:24	
Control Test		0.081	03:25	
Air Blank		0.000	03:25	
Diagnostics Check		OK	03:25	

Data Download Date/Time: 04/14/2017 03:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000955

INSTRUMENT REGISTERED TO: CLEWISTON P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: WILLIAMS,N,

BREATH TEST OPERATOR AGENCY: Clewiston PD

ARRESTING OFFICER: WILLIAMS,N

ARRESTING OFFICER AGENCY: Clewiston PD

LAST AGENCY INSPECTION: 03/01/2017

Date or Time Changed: No

DRY GAS STANDARD LOT #: 644476

DRY GAS EXPIRY DATE: 12/01/2019

VIOLATION CODE: DUI

ARREST TIME: 02:00

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:02

SUBJECT'S LAST NAME: HORVATH

DATE/TIME OF TEST: 03/04/2017 02:40:51

Diagnostics Check	OK	02:38	
Air Blank	0.000	02:38	
Control Test	0.080	02:38	
Air Blank	0.000	02:39	
Subject Sample	# 1	02:39	Subject Test Refused
Breath Volume	0.000	02:39	
Air Blank	0.000	02:39	
Control Test	0.081	02:40	
Air Blank	0.000	02:40	
Diagnostics Check	OK	02:40	

Data Download Date/Time: 04/14/2017 03:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000955

INSTRUMENT REGISTERED TO: CLEWISTON P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHAMBERLAIN,STEVEN,M
BREATH TEST OPERATOR AGENCY: Clewiston PD
ARRESTING OFFICER: TAYLOR,ANDREW
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 03/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 20:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:13

SUBJECT'S LAST NAME: SAUNDERS

DATE/TIME OF TEST: 03/10/2017 21:59:31

Diagnostics Check	OK	21:53
Air Blank	0.000	21:53
Control Test	0.080	21:54
Air Blank	0.000	21:54
Subject Sample # 1	0.038	21:55
Breath Volume	1.753	21:55
Air Blank	0.000	21:55
Air Blank	0.000	21:57
Subject Sample # 2	0.037	21:58
Breath Volume	1.640	21:58
Air Blank	0.000	21:58
Control Test	0.080	21:58
Air Blank	0.000	21:59
Diagnostics Check	OK	21:59

Data Download Date/Time: 04/14/2017 03:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000955

INSTRUMENT REGISTERED TO: CLEWISTON P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILIAMS,N,
BREATH TEST OPERATOR AGENCY: Clewiston PD
ARRESTING OFFICER: TAYLOR,A
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 03/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 09:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:57

SUBJECT'S LAST NAME: MOORE

DATE/TIME OF TEST: 03/19/2017 10:39:13

Diagnostics Check	OK	10:33
Air Blank	0.000	10:33
Control Test	0.081	10:33
Air Blank	0.000	10:34
Subject Sample # 1	0.023	10:34
Breath Volume	1.625	10:34
Air Blank	0.000	10:35
Air Blank	0.000	10:37
Subject Sample # 2	0.023	10:37
Breath Volume	1.367	10:37
Air Blank	0.000	10:38
Control Test	0.079	10:38
Air Blank	0.000	10:39
Diagnostics Check	OK	10:39

Data Download Date/Time: 04/14/2017 03:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000956

INSTRUMENT REGISTERED TO: JEFFERSON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCLELLAN,DETRICK,S
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: CAMPER,ALAN
ARRESTING OFFICER AGENCY: Jefferson County SO
LAST AGENCY INSPECTION: 02/14/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: VILORIO

DATE/TIME OF TEST: 03/06/2017 01:58:15

Diagnostics Check	OK	01:50
Air Blank	0.000	01:51
Control Test	0.077	01:51
Air Blank	0.000	01:52
Subject Sample # 1	0.183	01:53
Breath Volume	2.714	01:53
Air Blank	0.000	01:54
Air Blank	0.000	01:55
Subject Sample # 2	0.173	01:56
Breath Volume	2.203	01:56
Air Blank	0.000	01:57
Control Test	0.079	01:57
Air Blank	0.000	01:58
Diagnostics Check	OK	01:58

Data Download Date/Time: 03/14/2017 10:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000956

INSTRUMENT REGISTERED TO: JEFFERSON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SWINDLE, CHARLES, M
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: SWINDLE, CHARLES
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 02/14/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:28

SUBJECT'S LAST NAME: BUTLER

DATE/TIME OF TEST: 03/07/2017 04:14:25

Diagnostics Check	OK	04:07
Air Blank	0.000	04:07
Control Test	0.078	04:08
Air Blank	0.000	04:08
Subject Sample # 1	0.000	04:10
Breath Volume	1.582	04:10
Air Blank	0.000	04:10
Air Blank	0.000	04:12
Subject Sample # 2	0.000	04:13
Breath Volume	2.472	04:13
Air Blank	0.000	04:13
Control Test	0.079	04:13
Air Blank	0.000	04:14
Diagnostics Check	OK	04:14

Data Download Date/Time: 03/14/2017 10:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000956

INSTRUMENT REGISTERED TO: JEFFERSON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG, WILLIAM, B
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: PITTS, JACK
ARRESTING OFFICER AGENCY: Monticello PD
LAST AGENCY INSPECTION: 03/14/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:55

SUBJECT'S LAST NAME: VALLADARES

DATE/TIME OF TEST: 03/23/2017 22:22:41

Diagnostics Check	OK	22:16
Air Blank	0.000	22:16
Control Test	0.077	22:16
Air Blank	0.000	22:17
Subject Sample # 1	0.264	22:18
Breath Volume	2.582	22:18
Air Blank	0.000	22:18
Air Blank	0.000	22:20
Subject Sample # 2	0.270	22:21
Breath Volume	3.207	22:21
Air Blank	0.000	22:21
Control Test	0.077	22:22
Air Blank	0.000	22:22
Diagnostics Check	OK	22:22

Data Download Date/Time: 04/04/2017 11:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000956

INSTRUMENT REGISTERED TO: JEFFERSON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRELL, WILLIAM, E
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: HARRELL, WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 03/14/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 13:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:24

SUBJECT'S LAST NAME: THEODOS

DATE/TIME OF TEST: 03/26/2017 15:51:16

Diagnostics Check	OK	15:44
Air Blank	0.000	15:45
Control Test	0.078	15:45
Air Blank	0.000	15:46
Subject Sample # 1	0.202	15:46
Breath Volume	1.265	15:46
Air Blank	0.000	15:47
Air Blank	0.000	15:49
Subject Sample # 2	0.195	15:49
Breath Volume	1.308	15:49
Air Blank	0.000	15:50
Control Test	0.077	15:50
Air Blank	0.000	15:51
Diagnostics Check	OK	15:51

Data Download Date/Time: 04/04/2017 11:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEGRAW,JACK,D
BREATH TEST OPERATOR AGENCY: FL State University
ARRESTING OFFICER: DEGRAW,JACK
ARRESTING OFFICER AGENCY: FL State University
LAST AGENCY INSPECTION: 02/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:26

SUBJECT'S LAST NAME: CHRISTOPHER

DATE/TIME OF TEST: 03/03/2017 01:56:57

Diagnostics Check	OK	01:48
Air Blank	0.000	01:49
Control Test	0.080	01:49
Air Blank	0.000	01:50
Subject Sample # 1	0.165	01:51
Breath Volume	2.015	01:51
Air Blank	0.000	01:52
Air Blank	0.000	01:54
Subject Sample # 2	0.178	01:55
Breath Volume	3.191	01:55
Air Blank	0.000	01:56
Control Test	0.077	01:56
Air Blank	0.000	01:56
Diagnostics Check	OK	01:56

Data Download Date/Time: 03/08/2017 09:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000962

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEMPLIN,RICHARD,P
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: TEMPLIN,RICHARD
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 03/03/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 16:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:07

SUBJECT'S LAST NAME: MCDANIEL

DATE/TIME OF TEST: 03/17/2017 16:45:31

Diagnostics Check	OK	16:33	
Air Blank	0.000	16:34	
Control Test	0.080	16:34	
Air Blank	0.000	16:34	
Subject Sample	# 1	16:38	No Sample Provided
Breath Volume	0.000	16:38	
Air Blank	0.000	16:38	
Air Blank	0.000	16:40	
Subject Sample	# 2	16:43	No Sample Provided
Breath Volume	0.000	16:43	
Air Blank	0.000	16:44	
Control Test	0.080	16:44	
Air Blank	0.000	16:44	
Diagnostics Check	Fail	16:45	Temperature Regulation Test
Air Blank	0.000	16:45	

Data Download Date/Time: 04/13/2017 09:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MICHAEL,C
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: COMPTON,ALLEN
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 02/14/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 10:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 12:18

SUBJECT'S LAST NAME: LUCCA

DATE/TIME OF TEST: 03/03/2017 12:50:45

Diagnostics Check	OK	12:42
Air Blank	0.000	12:43
Control Test	0.078	12:43
Air Blank	0.000	12:44
Subject Sample # 1	0.227	12:44
Breath Volume	1.355	12:44
Air Blank	0.000	12:45
Air Blank	0.000	12:47
Subject Sample # 2	0.225	12:49
Breath Volume	1.769	12:49
Air Blank	0.000	12:49
Control Test	0.078	12:50
Air Blank	0.000	12:50
Diagnostics Check	OK	12:50

Data Download Date/Time: 03/09/2017 15:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,COURTNEY,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BORZA,ELLORY
ARRESTING OFFICER AGENCY: Winter Garden PD
LAST AGENCY INSPECTION: 02/14/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:14

SUBJECT'S LAST NAME: MCKINNEY

DATE/TIME OF TEST: 03/04/2017 04:43:12

Diagnostics Check	OK	04:37
Air Blank	0.000	04:37
Control Test	0.079	04:37
Air Blank	0.000	04:38
Subject Sample # 1	0.152	04:38
Breath Volume	3.023	04:38
Air Blank	0.000	04:39
Air Blank	0.000	04:41
Subject Sample # 2	0.149	04:41
Breath Volume	1.960	04:41
Air Blank	0.000	04:42
Control Test	0.079	04:42
Air Blank	0.000	04:43
Diagnostics Check	OK	04:43

Data Download Date/Time: 03/09/2017 15:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: VARELA,JOSE
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 02/14/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 06:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:00

SUBJECT'S LAST NAME: FREITES CALZADILLA

DATE/TIME OF TEST: 03/05/2017 08:31:06

Diagnostics Check	OK	08:24
Air Blank	0.000	08:24
Control Test	0.079	08:24
Air Blank	0.000	08:25
Subject Sample # 1	0.092	08:26
Breath Volume	1.421	08:26
Air Blank	0.000	08:27
Air Blank	0.000	08:29
Subject Sample # 2	0.098	08:29
Breath Volume	1.550	08:29
Air Blank	0.000	08:30
Control Test	0.079	08:30
Air Blank	0.000	08:31
Diagnostics Check	OK	08:31

Data Download Date/Time: 03/09/2017 15:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,COURTNEY,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MERKS,ANDREW
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 02/14/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:19

SUBJECT'S LAST NAME: JALLOW

DATE/TIME OF TEST: 03/06/2017 03:49:52

Diagnostics Check	OK	03:43
Air Blank	0.000	03:43
Control Test	0.079	03:44
Air Blank	0.000	03:44
Subject Sample # 1	0.175	03:45
Breath Volume	1.460	03:45
Air Blank	0.000	03:45
Air Blank	0.000	03:47
Subject Sample # 2	0.171	03:48
Breath Volume	1.394	03:48
Air Blank	0.000	03:48
Control Test	0.079	03:49
Air Blank	0.000	03:49
Diagnostics Check	OK	03:49

Data Download Date/Time: 03/09/2017 15:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,COURTNEY,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ROCCHILD,MIGUEL
ARRESTING OFFICER AGENCY: Winter Park PD
LAST AGENCY INSPECTION: 02/14/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:46

SUBJECT'S LAST NAME: HUNT

DATE/TIME OF TEST: 03/07/2017 03:15:01

Diagnostics Check	OK	03:08
Air Blank	0.000	03:09
Control Test	0.079	03:09
Air Blank	0.000	03:10
Subject Sample # 1	0.149	03:10
Breath Volume	2.660	03:10
Air Blank	0.000	03:11
Air Blank	0.000	03:13
Subject Sample # 2	0.137	03:13
Breath Volume	1.367	03:13
Air Blank	0.000	03:14
Control Test	0.079	03:14
Air Blank	0.000	03:14
Diagnostics Check	OK	03:15

Data Download Date/Time: 03/09/2017 15:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CANER,OSVALDO,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ANDREWS,GARY
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 02/14/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: BADALL

DATE/TIME OF TEST: 03/18/2017 04:05:17

Diagnostics Check	OK	03:58
Air Blank	0.000	03:58
Control Test	0.080	03:59
Air Blank	0.000	03:59
Subject Sample # 1	0.127	04:00
Breath Volume	1.441	04:00
Air Blank	0.000	04:01
Air Blank	0.000	04:03
Subject Sample # 2	0.119	04:03
Breath Volume	1.308	04:03
Air Blank	0.000	04:04
Control Test	0.078	04:04
Air Blank	0.000	04:05
Diagnostics Check	OK	04:05

Data Download Date/Time: 03/22/2017 11:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ, MICHAEL, J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: WONG SHUE, ANTHONY
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 02/14/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 10:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:15

SUBJECT'S LAST NAME: WACHIRA

DATE/TIME OF TEST: 03/19/2017 10:46:47

Diagnostics Check	OK	10:40
Air Blank	0.000	10:41
Control Test	0.078	10:41
Air Blank	0.000	10:42
Subject Sample # 1	0.136	10:42
Breath Volume	3.710	10:42
Air Blank	0.000	10:43
Air Blank	0.000	10:45
Subject Sample # 2	0.135	10:45
Breath Volume	1.410	10:45
Air Blank	0.000	10:45
Control Test	0.078	10:46
Air Blank	0.000	10:46
Diagnostics Check	OK	10:46

Data Download Date/Time: 03/22/2017 11:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000965

INSTRUMENT REGISTERED TO: QUINCY PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SPANGLER,GARY,A

BREATH TEST OPERATOR AGENCY: FHP Troop H

ARRESTING OFFICER: SPANGLER,GARY

ARRESTING OFFICER AGENCY: FHP Troop H

LAST AGENCY INSPECTION: 02/13/2017

Date or Time Changed: No

DRY GAS STANDARD LOT #: 34015080A2

DRY GAS EXPIRY DATE: 01/05/2018

VIOLATION CODE: DUI

ARREST TIME: 22:17

GENDER: Male

OBSERVATION PERIOD BEGAN: 23:05

SUBJECT'S LAST NAME: MALDONADO

DATE/TIME OF TEST: 03/04/2017 23:34:35

Diagnostics Check	OK	23:27
Air Blank	0.000	23:28
Control Test	0.080	23:28
Air Blank	0.000	23:28
Subject Sample # 1	0.165	23:29
Breath Volume	1.140	23:29
Air Blank	0.000	23:30
Air Blank	0.000	23:31
Subject Sample # 2	0.165	23:32
Breath Volume	2.093	23:32
Air Blank	0.000	23:33
Control Test	0.081	23:33
Air Blank	0.000	23:34
Diagnostics Check	OK	23:34

Data Download Date/Time: 03/23/2017 11:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000965

INSTRUMENT REGISTERED TO: QUINCY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHRITENOUR,STEVEN,M
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: WHRITENOUR,STEVEN
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 02/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:34

SUBJECT'S LAST NAME: GOMEZ

DATE/TIME OF TEST: 03/05/2017 05:01:02

Diagnostics Check	OK	04:54
Air Blank	0.000	04:55
Control Test	0.080	04:55
Air Blank	0.000	04:56
Subject Sample # 1	0.189	04:56
Breath Volume	3.531	04:56
Air Blank	0.000	04:57
Air Blank	0.000	04:58
Subject Sample # 2	0.195	04:59
Breath Volume	3.300	04:59
Air Blank	0.000	05:00
Control Test	0.081	05:00
Air Blank	0.000	05:00
Diagnostics Check	OK	05:01

Data Download Date/Time: 03/23/2017 11:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000965

INSTRUMENT REGISTERED TO: QUINCY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARGE,SUSAN,M
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: BARGE,SUSAN
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 02/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 08:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 11:15

SUBJECT'S LAST NAME: MERCER

DATE/TIME OF TEST: 03/06/2017 11:46:27

Diagnostics Check	OK	11:39
Air Blank	0.000	11:39
Control Test	0.079	11:40
Air Blank	0.000	11:40
Subject Sample # 1	0.029	11:41
Breath Volume	1.738	11:41
Air Blank	0.000	11:41
Air Blank	0.000	11:43
Subject Sample # 2	0.029	11:44
Breath Volume	2.726	11:44
Air Blank	0.000	11:45
Control Test	0.080	11:45
Air Blank	0.000	11:46
Diagnostics Check	OK	11:46

Data Download Date/Time: 03/23/2017 11:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000965

INSTRUMENT REGISTERED TO: QUINCY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPANGLER,GARY,A
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: SPANGLER,GARY
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 02/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:37

SUBJECT'S LAST NAME: COPPIANO

DATE/TIME OF TEST: 03/09/2017 03:08:55

Diagnostics Check	OK	02:57	
Air Blank	0.000	02:57	
Control Test	0.079	02:58	
Air Blank	0.000	02:58	
Subject Sample	# 1	03:01	No Sample Provided
Breath Volume	0.000	03:01	
Air Blank	0.000	03:02	
Air Blank	0.000	03:04	
Subject Sample	# 2	03:07	No Sample Provided
Breath Volume	0.000	03:07	
Air Blank	0.000	03:08	
Control Test	0.078	03:08	
Air Blank	0.000	03:08	
Diagnostics Check	OK	03:08	

Data Download Date/Time: 03/23/2017 11:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000965

INSTRUMENT REGISTERED TO: QUINCY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURNETTE,JAVON,A
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: BURNETTE,JAVON
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 02/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:50

SUBJECT'S LAST NAME: MONTROY

DATE/TIME OF TEST: 03/10/2017 20:17:24

Diagnostics Check	OK	20:10
Air Blank	0.000	20:11
Control Test	0.079	20:11
Air Blank	0.000	20:11
Subject Sample # 1	0.088	20:12
Breath Volume	1.625	20:12
Air Blank	0.000	20:13
Air Blank	0.000	20:15
Subject Sample # 2	0.095	20:15
Breath Volume	2.246	20:15
Air Blank	0.000	20:16
Control Test	0.079	20:16
Air Blank	0.000	20:17
Diagnostics Check	OK	20:17

Data Download Date/Time: 03/23/2017 11:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000965

INSTRUMENT REGISTERED TO: QUINCY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHRITENOUR,STEVEN,M
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: WHRITENOUR,STEVEN
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 02/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:50

SUBJECT'S LAST NAME: SCOTT

DATE/TIME OF TEST: 03/12/2017 20:22:27

Diagnostics Check	OK	20:11	
Air Blank	0.000	20:11	
Control Test	0.079	20:11	
Air Blank	0.000	20:12	
Subject Sample	# 1	20:15	No Sample Provided
Breath Volume	0.000	20:15	
Air Blank	0.000	20:15	
Air Blank	0.000	20:17	
Subject Sample	# 2	20:21	No Sample Provided
Breath Volume	0.000	20:21	
Air Blank	0.000	20:21	
Control Test	0.080	20:21	
Air Blank	0.000	20:22	
Diagnostics Check	OK	20:22	

Data Download Date/Time: 03/23/2017 11:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000965

INSTRUMENT REGISTERED TO: QUINCY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COPELAND,MICHAEL,C
BREATH TEST OPERATOR AGENCY: Quincy PD
ARRESTING OFFICER: COPELAND,MICHAEL
ARRESTING OFFICER AGENCY: Quincy PD
LAST AGENCY INSPECTION: 02/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 05:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:53

SUBJECT'S LAST NAME: REED

DATE/TIME OF TEST: 03/19/2017 09:19:31

Diagnostics Check	OK	09:13
Air Blank	0.000	09:14
Control Test	0.079	09:14
Air Blank	0.000	09:14
Subject Sample # 1	0.000	09:15
Breath Volume	2.757	09:15
Air Blank	0.000	09:15
Air Blank	0.000	09:17
Subject Sample # 2	0.000	09:18
Breath Volume	1.785	09:18
Air Blank	0.000	09:18
Control Test	0.080	09:18
Air Blank	0.000	09:19
Diagnostics Check	OK	09:19

Data Download Date/Time: 03/23/2017 11:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000965

INSTRUMENT REGISTERED TO: QUINCY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GAINES,CRAIG,A
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: GAINES,CRAIG
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 02/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 17:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:05

SUBJECT'S LAST NAME: NIX

DATE/TIME OF TEST: 03/19/2017 19:37:18

Diagnostics Check	OK	19:29
Air Blank	0.000	19:30
Control Test	0.080	19:30
Air Blank	0.000	19:31
Subject Sample # 1	0.266	19:32
Breath Volume	1.734	19:32
Air Blank	0.000	19:33
Air Blank	0.000	19:34
Subject Sample # 2	0.258	19:35
Breath Volume	2.289	19:35
Air Blank	0.000	19:36
Control Test	0.080	19:36
Air Blank	0.000	19:37
Diagnostics Check	OK	19:37

Data Download Date/Time: 03/23/2017 11:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000965

INSTRUMENT REGISTERED TO: QUINCY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLEN,JAMES,R
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: ALLEN,JAMES
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 03/23/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 11:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 11:56

SUBJECT'S LAST NAME: RALEY

DATE/TIME OF TEST: 03/25/2017 12:23:41

Diagnostics Check	OK	12:16
Air Blank	0.000	12:17
Control Test	0.079	12:17
Air Blank	0.000	12:18
Subject Sample # 1	0.028	12:19
Breath Volume	2.976	12:19
Air Blank	0.000	12:19
Air Blank	0.000	12:21
Subject Sample # 2	0.027	12:22
Breath Volume	3.140	12:22
Air Blank	0.000	12:22
Control Test	0.078	12:23
Air Blank	0.000	12:23
Diagnostics Check	OK	12:23

Data Download Date/Time: 04/07/2017 10:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: WAGER, WILLIAM, W

BREATH TEST OPERATOR AGENCY: St. Cloud PD

ARRESTING OFFICER: FLINT, RYAN

ARRESTING OFFICER AGENCY: St. Cloud PD

LAST AGENCY INSPECTION: 02/18/2017

Date or Time Changed: No

DRY GAS STANDARD LOT #: 163437

DRY GAS EXPIRY DATE: 10/30/2018

VIOLATION CODE: DUI

ARREST TIME: 16:55

GENDER: Male

OBSERVATION PERIOD BEGAN: 17:42

SUBJECT'S LAST NAME: KULIBERT

DATE/TIME OF TEST: 03/12/2017 18:04:25

Diagnostics Check	OK	18:03	
Air Blank		18:03	Ambient Fail
Air Blank	0.000	18:04	

Data Download Date/Time: 03/27/2017 16:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAGER, WILLIAM, W
BREATH TEST OPERATOR AGENCY: St. Cloud PD
ARRESTING OFFICER: FLINT, RYAN
ARRESTING OFFICER AGENCY: St. Cloud PD
LAST AGENCY INSPECTION: 02/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 16:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:42

SUBJECT'S LAST NAME: KULIBERT

DATE/TIME OF TEST: 03/12/2017 18:19:24

Diagnostics Check	OK	18:09
Air Blank	0.000	18:09
Control Test	0.079	18:09
Air Blank	0.000	18:10
Subject Sample # 1	0.210	18:10
Breath Volume	1.339	18:10
Air Blank	0.000	18:11
Air Blank	0.000	18:13
Subject Sample # 2	0.187	18:14
Breath Volume	1.492	18:14
Air Blank	0.000	18:15
Air Blank	0.000	18:17
Subject Sample # 3	0.193	18:17
Breath Volume	1.765	18:17
Air Blank	0.000	18:18
Control Test	0.077	18:18
Air Blank	0.000	18:19
Diagnostics Check	OK	18:19

Data Download Date/Time: 03/27/2017 16:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: APOPKA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPBELL,ROBERT,J
BREATH TEST OPERATOR AGENCY: Apopka PD
ARRESTING OFFICER: VAN SCHAACK,JOEL
ARRESTING OFFICER AGENCY: Apopka PD
LAST AGENCY INSPECTION: 02/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:38

SUBJECT'S LAST NAME: POE

DATE/TIME OF TEST: 03/03/2017 00:04:11

Diagnostics Check	OK	00:02	
Air Blank	0.000	00:02	
Control Test	0.076	00:02	
Air Blank	0.000	00:03	
Subject Sample	# 1	00:03	RFI Detect
Breath Volume	0.000	00:03	
Air Blank		00:03	RFI Detect
Air Blank	0.000	00:04	

Data Download Date/Time: 04/26/2017 08:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: APOPKA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPBELL,ROBERT,J
BREATH TEST OPERATOR AGENCY: Apopka PD
ARRESTING OFFICER: VAN SCHAACK,JOEL
ARRESTING OFFICER AGENCY: Apopka PD
LAST AGENCY INSPECTION: 02/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:38

SUBJECT'S LAST NAME: POE

DATE/TIME OF TEST: 03/03/2017 00:11:05

Diagnostics Check	OK	00:05	
Air Blank	0.000	00:05	
Control Test	0.075	00:06	
Air Blank	0.000	00:06	
Subject Sample # 1	0.000	00:06	
Breath Volume	2.414	00:06	
Air Blank	0.000	00:07	
Air Blank	0.000	00:09	
Subject Sample # 2	0.000	00:09	
Breath Volume	2.253	00:09	
Air Blank	0.000	00:10	
Control Test	0.074	00:10	Control Outside Tolerance
Air Blank	0.000	00:11	

Data Download Date/Time: 04/26/2017 08:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: APOPKA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPBELL,ROBERT,J
BREATH TEST OPERATOR AGENCY: Apopka PD
ARRESTING OFFICER: VAN SCHAACK,JOEL
ARRESTING OFFICER AGENCY: Apopka PD
LAST AGENCY INSPECTION: 02/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:32

SUBJECT'S LAST NAME: MORRIS

DATE/TIME OF TEST: 03/04/2017 00:59:29

Diagnostics Check	OK	00:53
Air Blank	0.000	00:53
Control Test	0.075	00:54
Air Blank	0.000	00:54
Subject Sample # 1	0.000	00:55
Breath Volume	3.019	00:55
Air Blank	0.000	00:55
Air Blank	0.000	00:57
Subject Sample # 2	0.000	00:58
Breath Volume	3.031	00:58
Air Blank	0.000	00:58
Control Test	0.075	00:58
Air Blank	0.000	00:59
Diagnostics Check	OK	00:59

Data Download Date/Time: 04/26/2017 08:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: APOPKA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPBELL,ROBERT,J
BREATH TEST OPERATOR AGENCY: Apopka PD
ARRESTING OFFICER: SORRELLS,BRIAN
ARRESTING OFFICER AGENCY: Apopka PD
LAST AGENCY INSPECTION: 02/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 17:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:39

SUBJECT'S LAST NAME: HELTON

DATE/TIME OF TEST: 03/16/2017 19:48:27

Diagnostics Check	OK	19:42
Air Blank	0.000	19:43
Control Test	0.076	19:43
Air Blank	0.000	19:43
Subject Sample # 1	0.000	19:44
Breath Volume	2.207	19:44
Air Blank	0.000	19:44
Air Blank	0.000	19:46
Subject Sample # 2	0.000	19:47
Breath Volume	1.984	19:47
Air Blank	0.000	19:47
Control Test	0.076	19:47
Air Blank	0.000	19:48
Diagnostics Check	OK	19:48

Data Download Date/Time: 04/26/2017 08:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: APOPKA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLER,ASHLEY,B
BREATH TEST OPERATOR AGENCY: Apopka PD
ARRESTING OFFICER: ELLER,ASHLEY
ARRESTING OFFICER AGENCY: Apopka PD
LAST AGENCY INSPECTION: 02/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:55

SUBJECT'S LAST NAME: FRANCO-RESENDIZ

DATE/TIME OF TEST: 03/18/2017 23:26:04

Diagnostics Check	OK	23:19
Air Blank	0.000	23:20
Control Test	0.076	23:20
Air Blank	0.000	23:21
Subject Sample # 1	0.117	23:21
Breath Volume	1.585	23:21
Air Blank	0.000	23:22
Air Blank	0.000	23:24
Subject Sample # 2	0.110	23:24
Breath Volume	1.191	23:24
Air Blank	0.000	23:25
Control Test	0.075	23:25
Air Blank	0.000	23:25
Diagnostics Check	OK	23:26

Data Download Date/Time: 04/26/2017 08:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: APOPKA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPBELL,ROBERT,J
BREATH TEST OPERATOR AGENCY: Apopka PD
ARRESTING OFFICER: DEVANIE,SEAN
ARRESTING OFFICER AGENCY: Apopka PD
LAST AGENCY INSPECTION: 02/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:41

SUBJECT'S LAST NAME: CURTIS

DATE/TIME OF TEST: 03/22/2017 03:13:44

Diagnostics Check	OK	03:07
Air Blank	0.000	03:07
Control Test	0.075	03:07
Air Blank	0.000	03:08
Subject Sample # 1	0.093	03:09
Breath Volume	2.738	03:09
Air Blank	0.000	03:10
Air Blank	0.000	03:11
Subject Sample # 2	0.090	03:12
Breath Volume	2.613	03:12
Air Blank	0.000	03:12
Control Test	0.075	03:13
Air Blank	0.000	03:13
Diagnostics Check	OK	03:13

Data Download Date/Time: 04/26/2017 08:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: APOPKA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLER,ASHLEY,B
BREATH TEST OPERATOR AGENCY: Apopka PD
ARRESTING OFFICER: GILZEN,MARCUS
ARRESTING OFFICER AGENCY: Apopka PD
LAST AGENCY INSPECTION: 02/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:03

SUBJECT'S LAST NAME: SANCHEZ JR

DATE/TIME OF TEST: 03/25/2017 00:30:48

Diagnostics Check	OK	00:27	
Air Blank	0.000	00:27	
Control Test	0.076	00:27	
Air Blank	0.000	00:28	
Subject Sample	# 1	00:29	Subject Test Refused
Breath Volume	0.000	00:29	
Air Blank	0.000	00:29	
Control Test	0.076	00:30	
Air Blank	0.000	00:30	
Diagnostics Check	OK	00:30	

Data Download Date/Time: 04/26/2017 08:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: APOPKA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPBELL,ROBERT,J
BREATH TEST OPERATOR AGENCY: Apopka PD
ARRESTING OFFICER: DEVANIE,SEAN
ARRESTING OFFICER AGENCY: Apopka PD
LAST AGENCY INSPECTION: 02/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: VERA

DATE/TIME OF TEST: 03/25/2017 03:19:01

Diagnostics Check	OK	03:17	
Air Blank	0.000	03:18	
Control Test	0.076	03:18	
Air Blank		03:18	RFI Detect
Air Blank	0.000	03:19	

Data Download Date/Time: 04/26/2017 08:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: APOPKA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPBELL,ROBERT,J
BREATH TEST OPERATOR AGENCY: Apopka PD
ARRESTING OFFICER: DEVANIE,SEAN
ARRESTING OFFICER AGENCY: Apopka PD
LAST AGENCY INSPECTION: 02/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: VERA

DATE/TIME OF TEST: 03/25/2017 03:27:02

Diagnostics Check	OK	03:20
Air Blank	0.000	03:20
Control Test	0.075	03:21
Air Blank	0.000	03:21
Subject Sample # 1	0.113	03:22
Breath Volume	1.312	03:22
Air Blank	0.000	03:23
Air Blank	0.000	03:24
Subject Sample # 2	0.108	03:25
Breath Volume	3.300	03:25
Air Blank	0.000	03:26
Control Test	0.075	03:26
Air Blank	0.000	03:26
Diagnostics Check	OK	03:27

Data Download Date/Time: 04/26/2017 08:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: APOPKA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLER,ASHLEY,B
BREATH TEST OPERATOR AGENCY: Apopka PD
ARRESTING OFFICER: GILZEN,MARCUS
ARRESTING OFFICER AGENCY: Apopka PD
LAST AGENCY INSPECTION: 02/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:57

SUBJECT'S LAST NAME: GOMEZ-PEREZ

DATE/TIME OF TEST: 03/26/2017 03:31:20

Diagnostics Check	OK	03:21	
Air Blank	0.000	03:21	
Control Test	0.075	03:22	
Air Blank	0.000	03:22	
Subject Sample	# 1 0.243	03:24	
Breath Volume	2.699	03:24	
Air Blank	0.000	03:24	
Air Blank	0.000	03:26	
Subject Sample	# 2 0.248	03:29	
Breath Volume	1.890	03:29	
Air Blank	0.000	03:30	
Control Test	0.072	03:30	Control Outside Tolerance
Air Blank	0.000	03:31	

Data Download Date/Time: 04/26/2017 08:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: APOPKA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLER,ASHLEY,B
BREATH TEST OPERATOR AGENCY: Apopka PD
ARRESTING OFFICER: GILZEN,MARCUS
ARRESTING OFFICER AGENCY: Apopka PD
LAST AGENCY INSPECTION: 02/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:57

SUBJECT'S LAST NAME: GOMEZ-PEREZ

DATE/TIME OF TEST: 03/26/2017 03:43:49

Diagnostics Check	OK	03:33	
Air Blank	0.000	03:33	
Control Test	0.075	03:34	
Air Blank	0.000	03:34	
Subject Sample	# 1 0.250	03:37	
Breath Volume	1.410	03:37	
Air Blank	0.000	03:38	
Air Blank	0.000	03:39	
Subject Sample	# 2 0.250	03:41	
Breath Volume	1.269	03:41	
Air Blank	0.000	03:42	
Control Test	0.072	03:43	Control Outside Tolerance
Air Blank	0.000	03:43	

Data Download Date/Time: 04/26/2017 08:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: APOPKA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLER,ASHLEY,B
BREATH TEST OPERATOR AGENCY: Apopka PD
ARRESTING OFFICER: CINTRON,KRISTIN
ARRESTING OFFICER AGENCY: Apopka PD
LAST AGENCY INSPECTION: 02/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 06:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:43

SUBJECT'S LAST NAME: FORD

DATE/TIME OF TEST: 03/27/2017 07:17:43

Diagnostics Check	OK	07:11
Air Blank	0.000	07:12
Control Test	0.075	07:12
Air Blank	0.000	07:12
Subject Sample # 1	0.000	07:13
Breath Volume	2.492	07:13
Air Blank	0.000	07:13
Air Blank	0.000	07:15
Subject Sample # 2	0.000	07:16
Breath Volume	3.128	07:16
Air Blank	0.000	07:16
Control Test	0.076	07:17
Air Blank	0.000	07:17
Diagnostics Check	OK	07:17

Data Download Date/Time: 04/26/2017 08:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRUZ,AYLER,
BREATH TEST OPERATOR AGENCY: FHP Troop K
ARRESTING OFFICER: CRUZ,AYLER
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 02/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 17:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:45

SUBJECT'S LAST NAME: MCGARRY

DATE/TIME OF TEST: 03/06/2017 19:12:31

Diagnostics Check	OK	19:06
Air Blank	0.000	19:07
Control Test	0.080	19:07
Air Blank	0.000	19:07
Subject Sample # 1	0.000	19:08
Breath Volume	1.468	19:08
Air Blank	0.000	19:08
Air Blank	0.000	19:10
Subject Sample # 2	0.000	19:11
Breath Volume	1.292	19:11
Air Blank	0.000	19:11
Control Test	0.081	19:11
Air Blank	0.000	19:12
Diagnostics Check	OK	19:12

Data Download Date/Time: 03/27/2017 16:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARRETT,JASON,R
BREATH TEST OPERATOR AGENCY: St. Cloud PD
ARRESTING OFFICER: BARRETT,JASON
ARRESTING OFFICER AGENCY: St. Cloud PD
LAST AGENCY INSPECTION: 02/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 09:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:20

SUBJECT'S LAST NAME: KENDLE

DATE/TIME OF TEST: 03/08/2017 11:08:44

Diagnostics Check	OK	11:02
Air Blank	0.000	11:02
Control Test	0.080	11:03
Air Blank	0.000	11:03
Subject Sample # 1	0.000	11:04
Breath Volume	2.222	11:04
Air Blank	0.000	11:04
Air Blank	0.000	11:06
Subject Sample # 2	0.000	11:07
Breath Volume	2.234	11:07
Air Blank	0.000	11:07
Control Test	0.080	11:08
Air Blank	0.000	11:08
Diagnostics Check	OK	11:08

Data Download Date/Time: 03/27/2017 16:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPOS,MICHAEL,A
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: CAMPOS,MICHAEL
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 02/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 05:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:30

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 03/10/2017 06:57:33

Diagnostics Check	OK	06:51
Air Blank	0.000	06:51
Control Test	0.081	06:52
Air Blank	0.000	06:52
Subject Sample # 1	0.109	06:52
Breath Volume	1.539	06:52
Air Blank	0.000	06:53
Air Blank	0.000	06:55
Subject Sample # 2	0.116	06:55
Breath Volume	1.867	06:55
Air Blank	0.000	06:56
Control Test	0.080	06:56
Air Blank	0.000	06:57
Diagnostics Check	OK	06:57

Data Download Date/Time: 03/27/2017 16:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEDULLA,VINCENT,A
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: PEDULLA,VINCENT
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 02/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 02:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: ROMAN

DATE/TIME OF TEST: 03/13/2017 02:43:56

Diagnostics Check		OK	02:37
Air Blank		0.000	02:38
Control Test		0.080	02:38
Air Blank		0.000	02:38
Subject Sample	# 1	0.210	02:39
Breath Volume		2.417	02:39
Air Blank		0.000	02:40
Air Blank		0.000	02:41
Subject Sample	# 2	0.204	02:42
Breath Volume		2.402	02:42
Air Blank		0.000	02:43
Control Test		0.080	02:43
Air Blank		0.000	02:43
Diagnostics Check		OK	02:43

Data Download Date/Time: 03/27/2017 16:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURROWS,STEPHEN,W
BREATH TEST OPERATOR AGENCY: St. Cloud PD
ARRESTING OFFICER: REDDITT,M
ARRESTING OFFICER AGENCY: St. Cloud PD
LAST AGENCY INSPECTION: 02/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: INGEBRITSON

DATE/TIME OF TEST: 03/13/2017 03:37:52

Diagnostics Check	OK	03:31
Air Blank	0.000	03:31
Control Test	0.080	03:31
Air Blank	0.000	03:32
Subject Sample # 1	0.157	03:33
Breath Volume	3.859	03:33
Air Blank	0.000	03:33
Air Blank	0.000	03:35
Subject Sample # 2	0.157	03:36
Breath Volume	4.394	03:36
Air Blank	0.000	03:36
Control Test	0.079	03:37
Air Blank	0.000	03:37
Diagnostics Check	OK	03:37

Data Download Date/Time: 03/27/2017 16:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPOS,MICHAEL,A
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: CAMPOS,MICHAEL
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 02/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 01:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: BABBS

DATE/TIME OF TEST: 03/14/2017 02:06:30

Diagnostics Check	OK	02:00
Air Blank	0.000	02:00
Control Test	0.080	02:01
Air Blank	0.000	02:01
Subject Sample # 1	0.121	02:02
Breath Volume	2.343	02:02
Air Blank	0.000	02:02
Air Blank	0.000	02:04
Subject Sample # 2	0.121	02:04
Breath Volume	1.777	02:04
Air Blank	0.000	02:05
Control Test	0.080	02:05
Air Blank	0.000	02:06
Diagnostics Check	OK	02:06

Data Download Date/Time: 03/27/2017 16:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LYONS,D,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: BEEKMAN,JA.
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 02/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 00:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:39

SUBJECT'S LAST NAME: PALLADINO

DATE/TIME OF TEST: 03/16/2017 01:06:31

Diagnostics Check	OK	01:00
Air Blank	0.000	01:00
Control Test	0.081	01:00
Air Blank	0.000	01:01
Subject Sample # 1	0.068	01:02
Breath Volume	2.222	01:02
Air Blank	0.000	01:02
Air Blank	0.000	01:04
Subject Sample # 2	0.065	01:05
Breath Volume	1.230	01:05
Air Blank	0.000	01:05
Control Test	0.081	01:05
Air Blank	0.000	01:06
Diagnostics Check	OK	01:06

Data Download Date/Time: 03/27/2017 16:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BEEKMAN,JACOB,C
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: BEEKMAN,JACOB
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 02/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 22:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:56

SUBJECT'S LAST NAME: GANESH

DATE/TIME OF TEST: 03/17/2017 23:24:20

Diagnostics Check	OK	23:16
Air Blank	0.000	23:17
Control Test	0.080	23:17
Air Blank	0.000	23:17
Subject Sample # 1	0.155	23:19
Breath Volume	3.066	23:19
Air Blank	0.000	23:20
Air Blank	0.000	23:21
Subject Sample # 2	0.154	23:22
Breath Volume	3.332	23:22
Air Blank	0.000	23:23
Control Test	0.080	23:23
Air Blank	0.000	23:24
Diagnostics Check	OK	23:24

Data Download Date/Time: 03/27/2017 16:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORUJO,PEDRO,L
BREATH TEST OPERATOR AGENCY: Kissimmee PD
ARRESTING OFFICER: CORUJO,PEDRO
ARRESTING OFFICER AGENCY: Kissimmee PD
LAST AGENCY INSPECTION: 02/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 00:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:13

SUBJECT'S LAST NAME: FLORES

DATE/TIME OF TEST: 03/18/2017 00:43:16

Diagnostics Check	OK	00:39	
Air Blank	0.000	00:40	
Control Test	0.080	00:40	
Air Blank	0.000	00:40	
Subject Sample	# 1 0.214	00:42	
Breath Volume	2.953	00:42	
Air Blank		00:42	Purge Fail
Air Blank	0.000	00:43	

Data Download Date/Time: 03/27/2017 16:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORUJO,PEDRO,L
BREATH TEST OPERATOR AGENCY: Kissimmee PD
ARRESTING OFFICER: CORUJO,PEDRO
ARRESTING OFFICER AGENCY: Kissimmee PD
LAST AGENCY INSPECTION: 02/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 00:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:13

SUBJECT'S LAST NAME: FLORES

DATE/TIME OF TEST: 03/18/2017 00:52:10

Diagnostics Check	OK	00:45
Air Blank	0.000	00:45
Control Test	0.080	00:45
Air Blank	0.000	00:46
Subject Sample # 1	0.201	00:46
Breath Volume	1.882	00:46
Air Blank	0.000	00:47
Air Blank	0.000	00:49
Subject Sample # 2	0.197	00:50
Breath Volume	2.375	00:50
Air Blank	0.000	00:51
Control Test	0.080	00:51
Air Blank	0.000	00:52
Diagnostics Check	OK	00:52

Data Download Date/Time: 03/27/2017 16:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEDULLA,VINCENT,A
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: PEDULLA,VINCENT
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 02/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 00:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:38

SUBJECT'S LAST NAME: HUTCHENS

DATE/TIME OF TEST: 03/18/2017 01:30:37

Diagnostics Check	OK	01:23
Air Blank	0.000	01:24
Control Test	0.080	01:24
Air Blank	0.000	01:25
Subject Sample # 1	0.179	01:26
Breath Volume	4.500	01:26
Air Blank	0.000	01:26
Air Blank	0.000	01:28
Subject Sample # 2	0.183	01:28
Breath Volume	3.597	01:28
Air Blank	0.000	01:29
Control Test	0.079	01:30
Air Blank	0.000	01:30
Diagnostics Check	OK	01:30

Data Download Date/Time: 03/27/2017 16:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPOS,MICHAEL,A
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: CAMPOS,MICHAEL
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 02/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 23:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:27

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 03/20/2017 01:00:39

Diagnostics Check	OK	00:48	
Air Blank	0.000	00:48	
Control Test	0.080	00:49	
Air Blank	0.000	00:49	
Subject Sample	# 1 0.200	00:50	
Breath Volume	1.332	00:50	
Air Blank	0.000	00:51	
Air Blank	0.000	00:52	
Subject Sample	# 2 0.196	00:56	Volume Not Met
Breath Volume	0.699	00:56	
Air Blank	0.000	00:56	
Air Blank	0.000	00:58	
Subject Sample	# 3 0.185	00:58	
Breath Volume	1.289	00:58	
Air Blank	0.000	00:59	
Control Test	0.078	01:00	
Air Blank	0.000	01:00	
Diagnostics Check	OK	01:00	

Data Download Date/Time: 03/27/2017 16:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHEINER,JAMES,R
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: SCHEINER,JAMES
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 02/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 36314080A3
DRY GAS EXPIRY DATE: 02/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:22

SUBJECT'S LAST NAME: COOK

DATE/TIME OF TEST: 03/07/2017 01:55:51

Diagnostics Check	OK	01:49
Air Blank	0.000	01:50
Control Test	0.079	01:50
Air Blank	0.000	01:51
Subject Sample # 1	0.094	01:51
Breath Volume	2.082	01:51
Air Blank	0.000	01:52
Air Blank	0.000	01:54
Subject Sample # 2	0.099	01:54
Breath Volume	2.476	01:54
Air Blank	0.000	01:54
Control Test	0.079	01:55
Air Blank	0.000	01:55
Diagnostics Check	OK	01:55

Data Download Date/Time: 03/17/2017 09:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHEINER,JAMES,R
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: VOLK,NICK
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 03/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 36314080A3
DRY GAS EXPIRY DATE: 02/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:09

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 03/22/2017 00:42:20

Diagnostics Check		OK	00:35
Air Blank		0.000	00:35
Control Test		0.076	00:35
Air Blank		0.000	00:36
Subject Sample	# 1	0.179	00:37
Breath Volume		3.062	00:37
Air Blank		0.000	00:37
Air Blank		0.000	00:39
Subject Sample	# 2	0.183	00:40
Breath Volume		3.226	00:40
Air Blank		0.000	00:41
Control Test		0.078	00:41
Air Blank		0.000	00:42
Diagnostics Check		OK	00:42

Data Download Date/Time: 05/04/2017 07:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WADE,CHARLES,M
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: STEVENSON,K
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 03/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 36314080A3
DRY GAS EXPIRY DATE: 02/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: WADE

DATE/TIME OF TEST: 03/26/2017 00:08:24

Diagnostics Check	OK	23:58
Air Blank	0.000	23:59
Control Test	0.079	23:59
Air Blank	0.000	00:00
Subject Sample # 1	0.152	00:03
Breath Volume	1.425	00:03
Air Blank	0.000	00:04
Air Blank	0.000	00:05
Subject Sample # 2	0.151	00:06
Breath Volume	1.609	00:06
Air Blank	0.000	00:07
Control Test	0.077	00:07
Air Blank	0.000	00:08
Diagnostics Check	OK	00:08

Data Download Date/Time: 05/04/2017 07:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

INSTRUMENT REGISTERED TO: TITUSVILLE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCGRORY,DIARMUID,J
BREATH TEST OPERATOR AGENCY: Titusville PD
ARRESTING OFFICER: MCGRORY,DIARMUID
ARRESTING OFFICER AGENCY: Titusville PD
LAST AGENCY INSPECTION: 03/07/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: BLANK

DATE/TIME OF TEST: 03/09/2017 03:13:56

Diagnostics Check	OK	03:07
Air Blank	0.000	03:08
Control Test	0.079	03:08
Air Blank	0.000	03:08
Subject Sample # 1	0.104	03:09
Breath Volume	2.371	03:09
Air Blank	0.000	03:09
Air Blank	0.000	03:11
Subject Sample # 2	0.104	03:12
Breath Volume	2.902	03:12
Air Blank	0.000	03:13
Control Test	0.080	03:13
Air Blank	0.000	03:13
Diagnostics Check	OK	03:13

Data Download Date/Time: 04/04/2017 11:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

INSTRUMENT REGISTERED TO: TITUSVILLE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCGRORY,DIARMUID,J
BREATH TEST OPERATOR AGENCY: Titusville PD
ARRESTING OFFICER: LINDO,MALIK
ARRESTING OFFICER AGENCY: Titusville PD
LAST AGENCY INSPECTION: 03/07/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:59

SUBJECT'S LAST NAME: KEELING

DATE/TIME OF TEST: 03/23/2017 21:27:55

Diagnostics Check	OK	21:20
Air Blank	0.000	21:20
Control Test	0.079	21:20
Air Blank	0.000	21:21
Subject Sample # 1	0.230	21:22
Breath Volume	1.109	21:22
Air Blank	0.000	21:23
Air Blank	0.000	21:25
Subject Sample # 2	0.220	21:26
Breath Volume	1.367	21:26
Air Blank	0.000	21:27
Control Test	0.079	21:27
Air Blank	0.000	21:27
Diagnostics Check	OK	21:27

Data Download Date/Time: 04/04/2017 11:54