Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001275

INSTRUMENT REGISTERED TO: HIALEAH GARDENS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,GONZALO,
BREATH TEST OPERATOR AGENCY: Medley PD
ARRESTING OFFICER: GARCIA,GONZALO
ARRESTING OFFICER AGENCY: Medley PD
LAST AGENCY INSPECTION: 05/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 08:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:10

SUBJECT'S LAST NAME: BONACHEA

DATE/TIME OF TEST: 06/10/2017 09:52:18

Diagnostics Check		OK	09:49	
Air Blank		0.000	09:49	
Control Test		0.077	09:50	
Air Blank		0.000	09:50	
Subject Sample	# 1		09:50	Subject Test Refused
Breath Volume		0.000	09:50	
Air Blank		0.000	09:51	
Control Test		0.078	09:51	
Air Blank		0.000	09:52	
Diagnostics Check		OK	09:52	

Data Download Date/Time: 06/22/2017 15:10

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYRTIL,REGINALD,J
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: FRANCO,GREGORY
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 06/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:42

SUBJECT'S LAST NAME: BAILEY

DATE/TIME OF TEST: 06/24/2017 05:28:36

Diagnostics Check		OK	05:17	
Air Blank		0.000	05:17	
Control Test		0.078	05:17	
Air Blank		0.000	05:18	
Subject Sample	# 1		05:21	No Sample Provided
Breath Volume		0.000	05:21	
Air Blank		0.000	05:22	
Air Blank		0.000	05:24	
Subject Sample	# 2		05:27	No Sample Provided
Breath Volume		0.000	05:27	
Air Blank		0.000	05:27	
Control Test		0.079	05:27	
Air Blank		0.000	05:28	
Diagnostics Check		OK	05:28	

Data Download Date/Time: 07/13/2017 11:16

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYRTIL,REGINALD,J
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: FRANCO,GREGORY
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 06/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:42

SUBJECT'S LAST NAME: BAILEY

DATE/TIME OF TEST: 06/24/2017 05:42:31

Diagnostics Check		OK	05:39	
Air Blank		0.000	05:40	
Control Test		0.079	05:40	
Air Blank		0.000	05:40	
Subject Sample	# 1		05:41	Subject Test Refused
Breath Volume		0.000	05:41	
Air Blank		0.000	05:41	
Control Test		0.079	05:41	
Air Blank		0.000	05:42	
Diagnostics Check		OK	05:42	

Data Download Date/Time: 07/13/2017 11:16

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRANCO,GREGORY,R
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: FRANCO,GREGORY
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 06/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:42

SUBJECT'S LAST NAME: BAILEY

DATE/TIME OF TEST: 06/24/2017 06:16:06

Diagnostics Check		OK	06:06
Air Blank		0.000	06:06
Control Test		0.080	06:07
Air Blank		0.000	06:07
Subject Sample	# 1	0.136	06:09
Breath Volume		1.226	06:09
Air Blank		0.000	06:10
Air Blank		0.000	06:12
Subject Sample	# 2	0.141	06:14
Breath Volume		1.312	06:14
Air Blank		0.000	06:15
Control Test		0.078	06:15
Air Blank		0.000	06:16
Diagnostics Check		OK	06:16

Data Download Date/Time: 07/13/2017 11:16

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DORISMOND,WISNER,
BREATH TEST OPERATOR AGENCY: Hialeah Gardens PD
ARRESTING OFFICER: DORISMOND,WISNER
ARRESTING OFFICER AGENCY: Hialeah Gardens PD
LAST AGENCY INSPECTION: 06/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 06/25/2017 02:28:13

Diagnostics Check		OK	02:20
Air Blank		0.000	02:21
Control Test		0.079	02:21
Air Blank		0.000	02:22
Subject Sample	# 1	0.138	02:22
Breath Volume		2.093	02:22
Air Blank		0.000	02:23
Air Blank		0.000	02:25
Subject Sample	# 2	0.133	02:26
Breath Volume		1.289	02:26
Air Blank		0.000	02:27
Control Test		0.078	02:27
Air Blank		0.000	02:28
Diagnostics Check		OK	02:28

Data Download Date/Time: 07/13/2017 11:16

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUERTA,EDWIN,
BREATH TEST OPERATOR AGENCY: Hialeah Gardens PD
ARRESTING OFFICER: HUERTA,EDWIN
ARRESTING OFFICER AGENCY: Hialeah Gardens PD
LAST AGENCY INSPECTION: 06/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: JAURENA

DATE/TIME OF TEST: 06/25/2017 02:41:48

Diagnostics Check		OK	02:38	
Air Blank		0.000	02:39	
Control Test		0.079	02:39	
Air Blank		0.000	02:40	
Subject Sample	# 1		02:40	Subject Test Refused
Breath Volume		0.000	02:40	
Air Blank		0.000	02:40	
Control Test		0.079	02:41	
Air Blank		0.000	02:41	
Diagnostics Check		OK	02:41	

Data Download Date/Time: 07/13/2017 11:16

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOROTIZA,JESUS,
BREATH TEST OPERATOR AGENCY: Hialeah Gardens PD
ARRESTING OFFICER: GOROTIZA,JESUS
ARRESTING OFFICER AGENCY: Hialeah Gardens PD
LAST AGENCY INSPECTION: 06/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: CRUZ PEREZ

DATE/TIME OF TEST: 06/25/2017 04:09:35

Diagnostics Check		OK	04:00
Air Blank		0.000	04:00
Control Test		0.080	04:01
Air Blank		0.000	04:01
Subject Sample	# 1	0.199	04:04
Breath Volume		2.960	04:04
Air Blank		0.000	04:05
Air Blank		0.000	04:07
Subject Sample	# 2	0.204	04:07
Breath Volume		3.218	04:07
Air Blank		0.000	04:08
Control Test		0.079	04:08
Air Blank		0.000	04:09
Diagnostics Check		OK	04:09

Data Download Date/Time: 07/13/2017 11:16

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 06/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:43

SUBJECT'S LAST NAME: NUNEZ

DATE/TIME OF TEST: 06/29/2017 04:10:43

Diagnostics Check		OK	04:07	
Air Blank		0.000	04:08	
Control Test		0.079	04:08	
Air Blank		0.000	04:09	
Subject Sample	# 1		04:09	Subject Test Refused
Breath Volume		0.000	04:09	
Air Blank		0.000	04:09	
Control Test		0.079	04:10	
Air Blank		0.000	04:10	
Diagnostics Check		OK	04:10	

Data Download Date/Time: 07/13/2017 11:16

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 06/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:36

SUBJECT'S LAST NAME: SUAREZ

DATE/TIME OF TEST: 06/30/2017 05:11:36

Diagnostics Check		OK	05:08	
Air Blank		0.000	05:09	
Control Test		0.079	05:09	
Air Blank		0.000	05:09	
Subject Sample	# 1		05:10	Subject Test Refused
Breath Volume		0.000	05:10	
Air Blank		0.000	05:10	
Control Test		0.079	05:10	
Air Blank		0.000	05:11	
Diagnostics Check		OK	05:11	

Data Download Date/Time: 07/13/2017 11:16

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MOSLEY,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:27

SUBJECT'S LAST NAME: GALLO FLORES

DATE/TIME OF TEST: 06/01/2017 01:50:14

Diagnostics Check		OK	01:47	
Air Blank		0.000	01:47	
Control Test		0.080	01:48	
Air Blank		0.000	01:48	
Subject Sample	# 1		01:48	Subject Test Refused
Breath Volume		0.000	01:48	
Air Blank		0.000	01:49	
Control Test		0.080	01:49	
Air Blank		0.000	01:50	
Diagnostics Check		OK	01:50	

Data Download Date/Time: 06/15/2017 12:41

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: RODRIGUEZ,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: KENNISON-NOVAK

DATE/TIME OF TEST: 06/01/2017 04:21:23

Diagnostics Check		OK	04:15
Air Blank		0.000	04:15
Control Test		0.080	04:16
Air Blank		0.000	04:16
Subject Sample	# 1	0.027	04:17
Breath Volume		5.652	04:17
Air Blank		0.000	04:17
Air Blank		0.000	04:19
Subject Sample	# 2	0.024	04:19
Breath Volume		5.164	04:19
Air Blank		0.000	04:20
Control Test		0.079	04:20
Air Blank		0.000	04:21
Diagnostics Check		OK	04:21

Data Download Date/Time: 06/15/2017 12:42

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: KERSHAW,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:10

SUBJECT'S LAST NAME: LOISEL

DATE/TIME OF TEST: 06/03/2017 05:37:00

Diagnostics Check		OK	05:30
Air Blank		0.000	05:30
Control Test		0.079	05:30
Air Blank		0.000	05:31
Subject Sample	# 1	0.220	05:32
Breath Volume		2.546	05:32
Air Blank		0.000	05:33
Air Blank		0.000	05:34
Subject Sample	# 2	0.228	05:35
Breath Volume		2.464	05:35
Air Blank		0.000	05:36
Control Test		0.079	05:36
Air Blank		0.000	05:36
Diagnostics Check		OK	05:37

Data Download Date/Time: 06/15/2017 12:42

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MOORE,A
ARRESTING OFFICER AGENCY: Atlantic Beach PD
LAST AGENCY INSPECTION: 05/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:15

SUBJECT'S LAST NAME: TORRES GARCIA

DATE/TIME OF TEST: 06/04/2017 04:43:43

Diagnostics Check		OK	04:35
Air Blank		0.000	04:35
Control Test		0.080	04:35
Air Blank		0.000	04:36
Subject Sample	# 1	0.153	04:38
Breath Volume		1.742	04:38
Air Blank		0.000	04:39
Air Blank		0.000	04:40
Subject Sample	# 2	0.157	04:42
Breath Volume		1.855	04:42
Air Blank		0.000	04:42
Control Test		0.079	04:43
Air Blank		0.000	04:43
Diagnostics Check		OK	04:43

Data Download Date/Time: 06/15/2017 12:42

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,F.,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BORGERT,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 08:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:25

SUBJECT'S LAST NAME: HARTLEY

DATE/TIME OF TEST: 06/04/2017 08:51:24

Diagnostics Check		OK	08:45
Air Blank		0.000	08:45
Control Test		0.080	08:46
Air Blank		0.000	08:46
Subject Sample	# 1	0.187	08:47
Breath Volume		2.449	08:47
Air Blank		0.000	08:47
Air Blank		0.000	08:49
Subject Sample	# 2	0.194	08:49
Breath Volume		2.359	08:49
Air Blank		0.000	08:50
Control Test		0.079	08:50
Air Blank		0.000	08:51
Diagnostics Check		OK	08:51

Data Download Date/Time: 06/15/2017 12:42

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,F.,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ANTONINI,M
ARRESTING OFFICER AGENCY: Atlantic Beach PD
LAST AGENCY INSPECTION: 05/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 09:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:48

SUBJECT'S LAST NAME: MEEKINS

DATE/TIME OF TEST: 06/04/2017 10:25:27

Diagnostics Check		OK	10:11	
Air Blank		0.000	10:11	
Control Test		0.080	10:12	
Air Blank		0.000	10:12	
Subject Sample	# 1	0.143	10:13	
Breath Volume		1.125	10:13	
Air Blank		0.000	10:13	
Air Blank		0.000	10:15	
Subject Sample	# 2	0.116	10:18	Volume Not Met
Breath Volume		0.773	10:18	
Air Blank		0.000	10:19	
Air Blank		0.000	10:21	
Subject Sample	# 3	0.135	10:23	
Breath Volume		1.441	10:23	
Air Blank		0.000	10:24	
Control Test		0.079	10:24	
Air Blank		0.000	10:25	
Diagnostics Check		OK	10:25	

Data Download Date/Time: 06/15/2017 12:42

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FILBERT,E
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 05/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:29

SUBJECT'S LAST NAME: COMEAU

DATE/TIME OF TEST: 06/07/2017 04:57:15

Diagnostics Check		OK	04:54	
Air Blank		0.000	04:54	
Control Test		0.080	04:55	
Air Blank		0.000	04:55	
Subject Sample	# 1		04:55	Subject Test Refused
Breath Volume		0.000	04:55	
Air Blank		0.000	04:56	
Control Test		0.079	04:56	
Air Blank		0.000	04:57	
Diagnostics Check		OK	04:57	

Data Download Date/Time: 06/15/2017 12:42

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHITE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: CARPENTER

DATE/TIME OF TEST: 06/08/2017 02:45:43

Diagnostics Check		OK	02:42	
Air Blank		0.000	02:43	
Control Test		0.080	02:43	
Air Blank		0.000	02:44	
Subject Sample	# 1		02:44	Subject Test Refused
Breath Volume		0.000	02:44	
Air Blank		0.000	02:44	
Control Test		0.080	02:45	
Air Blank		0.000	02:45	
Diagnostics Check		OK	02:45	

Data Download Date/Time: 06/15/2017 12:42

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MOSLEY,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:04

SUBJECT'S LAST NAME: MCCLOUD

DATE/TIME OF TEST: 06/08/2017 04:30:35

Diagnostics Check		OK	04:24
Air Blank		0.000	04:25
Control Test		0.080	04:25
Air Blank		0.000	04:26
Subject Sample	# 1	0.240	04:26
Breath Volume		1.562	04:26
Air Blank		0.000	04:26
Air Blank		0.000	04:28
Subject Sample	# 2	0.241	04:29
Breath Volume		1.339	04:29
Air Blank		0.000	04:29
Control Test		0.079	04:30
Air Blank		0.000	04:30
Diagnostics Check		OK	04:30

Data Download Date/Time: 06/15/2017 12:42

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPICER,LORETHA,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: TERRELL,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 14:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:40

SUBJECT'S LAST NAME: RAIFORD

DATE/TIME OF TEST: 06/08/2017 16:16:12

Diagnostics Check		OK	16:01	
Air Blank		0.000	16:01	
Control Test		0.080	16:01	
Air Blank		0.000	16:02	
Subject Sample	# 1	0.083	16:03	
Breath Volume		1.136	16:03	
Air Blank		0.000	16:04	
Air Blank		0.000	16:05	
Subject Sample	# 2	0.074	16:09	Volume Not Met
Breath Volume		1.468	16:09	
Air Blank		0.000	16:09	
Air Blank		0.000	16:11	
Subject Sample	# 3	0.068	16:14	Volume Not Met
Breath Volume		0.425	16:14	
Air Blank		0.000	16:15	
Control Test		0.080	16:15	
Air Blank		0.000	16:16	
Diagnostics Check		OK	16:16	

Data Download Date/Time: 06/15/2017 12:42

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPICER,LORETHA,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: STANLEY,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 18:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:58

SUBJECT'S LAST NAME: SPARKS-BING

DATE/TIME OF TEST: 06/08/2017 20:24:47

Diagnostics Check		OK	20:18
Air Blank		0.000	20:19
Control Test		0.079	20:19
Air Blank		0.000	20:20
Subject Sample	# 1	0.000	20:20
Breath Volume		2.664	20:20
Air Blank		0.000	20:21
Air Blank		0.000	20:23
Subject Sample	# 2	0.000	20:23
Breath Volume		2.644	20:23
Air Blank		0.000	20:23
Control Test		0.079	20:24
Air Blank		0.000	20:24
Diagnostics Check		OK	20:24

Data Download Date/Time: 06/15/2017 12:42

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHITE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:07

SUBJECT'S LAST NAME: MANICK

DATE/TIME OF TEST: 06/08/2017 23:33:46

Diagnostics Check		OK	23:27
Air Blank		0.000	23:27
Control Test		0.079	23:28
Air Blank		0.000	23:28
Subject Sample	# 1	0.148	23:29
Breath Volume		1.464	23:29
Air Blank		0.000	23:30
Air Blank		0.000	23:31
Subject Sample	# 2	0.144	23:32
Breath Volume		1.230	23:32
Air Blank		0.000	23:32
Control Test		0.079	23:33
Air Blank		0.000	23:33
Diagnostics Check		OK	23:33

Data Download Date/Time: 06/15/2017 12:42

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: STANLEY,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: FLOWERS

DATE/TIME OF TEST: 06/09/2017 02:17:01

Diagnostics Check		OK	02:10
Air Blank		0.000	02:10
Control Test		0.080	02:11
Air Blank		0.000	02:11
Subject Sample	# 1	0.162	02:12
Breath Volume		1.273	02:12
Air Blank		0.000	02:13
Air Blank		0.000	02:15
Subject Sample	# 2	0.171	02:15
Breath Volume		1.457	02:15
Air Blank		0.000	02:16
Control Test		0.079	02:16
Air Blank		0.000	02:16
Diagnostics Check		OK	02:17

Data Download Date/Time: 06/15/2017 12:42

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FARLEY,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: LACHUT

DATE/TIME OF TEST: 06/09/2017 03:16:53

Diagnostics Check		OK	03:10
Air Blank		0.000	03:11
Control Test		0.080	03:11
Air Blank		0.000	03:12
Subject Sample	# 1	0.156	03:12
Breath Volume		1.796	03:12
Air Blank		0.000	03:12
Air Blank		0.000	03:14
Subject Sample	# 2	0.155	03:15
Breath Volume		2.109	03:15
Air Blank		0.000	03:15
Control Test		0.080	03:16
Air Blank		0.000	03:16
Diagnostics Check		OK	03:16

Data Download Date/Time: 06/15/2017 12:42

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: STANLEY,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: FRIESON

DATE/TIME OF TEST: 06/10/2017 02:55:20

Diagnostics Check		OK	02:51	
Air Blank		0.000	02:52	
Control Test		0.080	02:52	
Air Blank		0.000	02:53	
Subject Sample	# 1		02:53	Subject Test Refused
Breath Volume		0.000	02:53	
Air Blank		0.000	02:54	
Control Test		0.079	02:54	
Air Blank		0.000	02:55	
Diagnostics Check		OK	02:55	

Data Download Date/Time: 06/15/2017 12:42

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DOUGHFMAN,D
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:13

SUBJECT'S LAST NAME: RICKER

DATE/TIME OF TEST: 06/10/2017 03:42:19

Diagnostics Check		OK	03:36
Air Blank		0.000	03:36
Control Test		0.079	03:37
Air Blank		0.000	03:37
Subject Sample	# 1	0.193	03:37
Breath Volume		1.625	03:37
Air Blank		0.000	03:38
Air Blank		0.000	03:40
Subject Sample	# 2	0.207	03:40
Breath Volume		1.425	03:40
Air Blank		0.000	03:41
Control Test		0.079	03:41
Air Blank		0.000	03:42
Diagnostics Check		OK	03:42

Data Download Date/Time: 06/15/2017 12:42

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHITE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:32

SUBJECT'S LAST NAME: CHAMPION

DATE/TIME OF TEST: 06/10/2017 04:59:03

Diagnostics Check		OK	04:53
Air Blank		0.000	04:53
Control Test		0.080	04:53
Air Blank		0.000	04:54
Subject Sample	# 1	0.209	04:54
Breath Volume		1.988	04:54
Air Blank		0.000	04:55
Air Blank		0.000	04:57
Subject Sample	# 2	0.208	04:57
Breath Volume		1.261	04:57
Air Blank		0.000	04:58
Control Test		0.079	04:58
Air Blank		0.000	04:58
Diagnostics Check		OK	04:59

Data Download Date/Time: 06/15/2017 12:42

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CAMACHO,K
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 05/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 05:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:10

SUBJECT'S LAST NAME: BURGOS CALDERON

DATE/TIME OF TEST: 06/10/2017 06:34:41

Diagnostics Check		OK	06:31	
Air Blank		0.000	06:32	
Control Test		0.079	06:32	
Air Blank		0.000	06:33	
Subject Sample	# 1		06:33	Subject Test Refused
Breath Volume		0.000	06:33	
Air Blank		0.000	06:33	
Control Test		0.080	06:34	
Air Blank		0.000	06:34	
Diagnostics Check		OK	06:34	

Data Download Date/Time: 06/15/2017 12:42

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BRYANT,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: PONDER

DATE/TIME OF TEST: 06/11/2017 03:02:32

Diagnostics Check		OK	02:55
Air Blank		0.000	02:55
Control Test		0.080	02:56
Air Blank		0.000	02:56
Subject Sample	# 1	0.214	02:57
Breath Volume		1.667	02:57
Air Blank		0.000	02:58
Air Blank		0.000	03:00
Subject Sample	# 2	0.210	03:01
Breath Volume		1.671	03:01
Air Blank		0.000	03:01
Control Test		0.078	03:01
Air Blank		0.000	03:02
Diagnostics Check		OK	03:02

Data Download Date/Time: 06/15/2017 12:42

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SIMMONS,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:06

SUBJECT'S LAST NAME: HAWTHORNE

DATE/TIME OF TEST: 06/11/2017 04:29:15

Diagnostics Check		OK	04:26	
Air Blank		0.000	04:26	
Control Test		0.080	04:27	
Air Blank		0.000	04:27	
Subject Sample	# 1		04:27	Subject Test Refused
Breath Volume		0.000	04:27	
Air Blank		0.000	04:28	
Control Test		0.079	04:28	
Air Blank		0.000	04:29	
Diagnostics Check		OK	04:29	

Data Download Date/Time: 06/15/2017 12:42

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MORRISON,M
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 05/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:53

SUBJECT'S LAST NAME: CRUZ

DATE/TIME OF TEST: 06/11/2017 05:18:32

Diagnostics Check		OK	05:15	
Air Blank		0.000	05:16	
Control Test		0.079	05:16	
Air Blank		0.000	05:16	
Subject Sample	# 1		05:17	Subject Test Refused
Breath Volume		0.000	05:17	
Air Blank		0.000	05:17	
Control Test		0.079	05:17	
Air Blank		0.000	05:18	
Diagnostics Check		OK	05:18	

Data Download Date/Time: 06/15/2017 12:42

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BROWN,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:06

SUBJECT'S LAST NAME: PERNELL

DATE/TIME OF TEST: 06/12/2017 02:34:03

Diagnostics Check		OK	02:27
Air Blank		0.000	02:27
Control Test		0.080	02:28
Air Blank		0.000	02:28
Subject Sample	# 1	0.149	02:29
Breath Volume		1.148	02:29
Air Blank		0.000	02:29
Air Blank		0.000	02:31
Subject Sample	# 2	0.144	02:32
Breath Volume		1.289	02:32
Air Blank		0.000	02:33
Control Test		0.080	02:33
Air Blank		0.000	02:33
Diagnostics Check		OK	02:34

Data Download Date/Time: 06/15/2017 12:42

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: KERSHAW,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:02

SUBJECT'S LAST NAME: ADAMS

DATE/TIME OF TEST: 06/12/2017 05:29:30

Diagnostics Check		OK	05:23
Air Blank		0.000	05:23
Control Test		0.080	05:23
Air Blank		0.000	05:24
Subject Sample	# 1	0.222	05:25
Breath Volume		1.750	05:25
Air Blank		0.000	05:25
Air Blank		0.000	05:27
Subject Sample	# 2	0.202	05:28
Breath Volume		2.218	05:28
Air Blank		0.000	05:28
Control Test		0.080	05:28
Air Blank		0.000	05:29
Diagnostics Check		OK	05:29

Data Download Date/Time: 06/15/2017 12:42

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BROWN,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:37

SUBJECT'S LAST NAME: FLETCHER

DATE/TIME OF TEST: 06/14/2017 03:04:46

Diagnostics Check		OK	02:58
Air Blank		0.000	02:58
Control Test		0.080	02:58
Air Blank		0.000	02:59
Subject Sample	# 1	0.068	02:59
Breath Volume		2.054	02:59
Air Blank		0.000	03:00
Air Blank		0.000	03:02
Subject Sample	# 2	0.067	03:03
Breath Volume		1.976	03:03
Air Blank		0.000	03:03
Control Test		0.079	03:04
Air Blank		0.000	03:04
Diagnostics Check		OK	03:04

Data Download Date/Time: 06/15/2017 12:42

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: TORRES,W
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 05/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:28

SUBJECT'S LAST NAME: HARPER

DATE/TIME OF TEST: 06/15/2017 03:55:10

Diagnostics Check		OK	03:48
Air Blank		0.000	03:49
Control Test		0.080	03:49
Air Blank		0.000	03:50
Subject Sample	# 1	0.102	03:50
Breath Volume		2.511	03:50
Air Blank		0.000	03:51
Air Blank		0.000	03:53
Subject Sample	# 2	0.098	03:53
Breath Volume		2.398	03:53
Air Blank		0.000	03:54
Control Test		0.080	03:54
Air Blank		0.000	03:55
Diagnostics Check		OK	03:55

Data Download Date/Time: 06/15/2017 12:42

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHITE,J'
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:36

SUBJECT'S LAST NAME: ROMERO

DATE/TIME OF TEST: 06/16/2017 03:01:14

Diagnostics Check		OK	02:58	
Air Blank		0.000	02:58	
Control Test		0.080	02:59	
Air Blank		0.000	02:59	
Subject Sample	# 1		02:59	Subject Test Refused
Breath Volume		0.000	02:59	
Air Blank		0.000	03:00	
Control Test		0.080	03:00	
Air Blank		0.000	03:01	
Diagnostics Check		OK	03:01	

Data Download Date/Time: 06/27/2017 14:38

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MOSLEY,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:04

SUBJECT'S LAST NAME: TROUTMAN

DATE/TIME OF TEST: 06/16/2017 04:31:59

Diagnostics Check		OK	04:24
Air Blank		0.000	04:25
Control Test		0.080	04:25
Air Blank		0.000	04:25
Subject Sample	# 1	0.218	04:26
Breath Volume		2.167	04:26
Air Blank		0.000	04:27
Air Blank		0.000	04:29
Subject Sample	# 2	0.203	04:30
Breath Volume		2.683	04:30
Air Blank		0.000	04:31
Control Test		0.080	04:31
Air Blank		0.000	04:31
Diagnostics Check		OK	04:31

Data Download Date/Time: 06/27/2017 14:38

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PADGETT,C,N
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HYINK,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 16:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:18

SUBJECT'S LAST NAME: WILKES JR

DATE/TIME OF TEST: 06/16/2017 17:48:28

Diagnostics Check		OK	17:39
Air Blank		0.000	17:39
Control Test		0.080	17:39
Air Blank		0.000	17:40
Subject Sample	# 1	0.111	17:42
Breath Volume		1.382	17:42
Air Blank		0.000	17:43
Air Blank		0.000	17:45
Subject Sample	# 2	0.106	17:46
Breath Volume		1.964	17:46
Air Blank		0.000	17:47
Control Test		0.079	17:47
Air Blank		0.000	17:48
Diagnostics Check		OK	17:48

Data Download Date/Time: 06/27/2017 14:38

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: RODRIGUEZ,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:56

SUBJECT'S LAST NAME: LANE

DATE/TIME OF TEST: 06/17/2017 00:23:37

Diagnostics Check		OK	00:16
Air Blank		0.000	00:16
Control Test		0.080	00:17
Air Blank		0.000	00:17
Subject Sample	# 1	0.163	00:18
Breath Volume		1.273	00:18
Air Blank		0.000	00:19
Air Blank		0.000	00:21
Subject Sample	# 2	0.166	00:22
Breath Volume		1.492	00:22
Air Blank		0.000	00:22
Control Test		0.079	00:23
Air Blank		0.000	00:23
Diagnostics Check		OK	00:23

Data Download Date/Time: 06/27/2017 14:38

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FILBERT,E
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 06/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:27

SUBJECT'S LAST NAME: WRIGHT

DATE/TIME OF TEST: 06/17/2017 01:54:20

Diagnostics Check		OK	01:47
Air Blank		0.000	01:47
Control Test		0.080	01:48
Air Blank		0.000	01:48
Subject Sample	# 1	0.129	01:49
Breath Volume		3.546	01:49
Air Blank		0.000	01:50
Air Blank		0.000	01:52
Subject Sample	# 2	0.127	01:52
Breath Volume		2.109	01:52
Air Blank		0.000	01:53
Control Test		0.080	01:53
Air Blank		0.000	01:54
Diagnostics Check		OK	01:54

Data Download Date/Time: 06/27/2017 14:38

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FARLEY,J
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 06/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: MORRIS

DATE/TIME OF TEST: 06/17/2017 02:39:31

Diagnostics Check		OK	02:36	
Air Blank		0.000	02:37	
Control Test		0.080	02:37	
Air Blank		0.000	02:38	
Subject Sample	# 1		02:38	Subject Test Refused
Breath Volume		0.000	02:38	
Air Blank		0.000	02:38	
Control Test		0.079	02:38	
Air Blank		0.000	02:39	
Diagnostics Check		OK	02:39	

Data Download Date/Time: 06/27/2017 14:38

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HOOVER,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: BAGAYOKO

DATE/TIME OF TEST: 06/17/2017 03:58:02

Diagnostics Check		OK	03:50
Air Blank		0.000	03:50
Control Test		0.080	03:51
Air Blank		0.000	03:51
Subject Sample	# 1	0.142	03:52
Breath Volume		1.429	03:52
Air Blank		0.000	03:53
Air Blank		0.000	03:55
Subject Sample	# 2	0.143	03:56
Breath Volume		1.558	03:56
Air Blank		0.000	03:57
Control Test		0.079	03:57
Air Blank		0.000	03:57
Diagnostics Check		OK	03:58

Data Download Date/Time: 06/27/2017 14:38

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HANSON,C
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 06/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:56

SUBJECT'S LAST NAME: SYLVIA

DATE/TIME OF TEST: 06/17/2017 05:23:02

Diagnostics Check		OK	05:16
Air Blank		0.000	05:17
Control Test		0.080	05:17
Air Blank		0.000	05:18
Subject Sample	# 1	0.125	05:18
Breath Volume		3.269	05:18
Air Blank		0.000	05:19
Air Blank		0.000	05:20
Subject Sample	# 2	0.117	05:21
Breath Volume		1.683	05:21
Air Blank		0.000	05:22
Control Test		0.079	05:22
Air Blank		0.000	05:22
Diagnostics Check		OK	05:23

Data Download Date/Time: 06/27/2017 14:38

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHITE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:01

SUBJECT'S LAST NAME: ROGERS

DATE/TIME OF TEST: 06/18/2017 00:28:29

Diagnostics Check		OK	00:21
Air Blank		0.000	00:22
Control Test		0.080	00:22
Air Blank		0.000	00:23
Subject Sample	# 1	0.219	00:23
Breath Volume		1.199	00:23
Air Blank		0.000	00:24
Air Blank		0.000	00:26
Subject Sample	# 2	0.226	00:26
Breath Volume		1.839	00:26
Air Blank		0.000	00:27
Control Test		0.080	00:27
Air Blank		0.000	00:28
Diagnostics Check		OK	00:28

Data Download Date/Time: 06/27/2017 14:38

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HOOVER,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:01

SUBJECT'S LAST NAME: BRITT

DATE/TIME OF TEST: 06/18/2017 03:27:47

Diagnostics Check		OK	03:21
Air Blank		0.000	03:22
Control Test		0.080	03:22
Air Blank		0.000	03:22
Subject Sample	# 1	0.171	03:23
Breath Volume		3.132	03:23
Air Blank		0.000	03:23
Air Blank		0.000	03:25
Subject Sample	# 2	0.167	03:26
Breath Volume		3.226	03:26
Air Blank		0.000	03:26
Control Test		0.080	03:27
Air Blank		0.000	03:27
Diagnostics Check		OK	03:27

Data Download Date/Time: 06/27/2017 14:38

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BRETT,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:02

SUBJECT'S LAST NAME: DENHAM

DATE/TIME OF TEST: 06/18/2017 05:28:13

Diagnostics Check		OK	05:22
Air Blank		0.000	05:22
Control Test		0.080	05:22
Air Blank		0.000	05:23
Subject Sample	# 1	0.117	05:23
Breath Volume		1.679	05:23
Air Blank		0.000	05:24
Air Blank		0.000	05:26
Subject Sample	# 2	0.122	05:26
Breath Volume		1.644	05:26
Air Blank		0.000	05:27
Control Test		0.079	05:27
Air Blank		0.000	05:28
Diagnostics Check		OK	05:28

Data Download Date/Time: 06/27/2017 14:38

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: GRUNWALD,K
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:56

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 06/18/2017 06:19:52

Diagnostics Check		OK	06:17	
Air Blank		0.000	06:17	
Control Test		0.080	06:17	
Air Blank		0.000	06:18	
Subject Sample	# 1		06:18	Subject Test Refused
Breath Volume		0.000	06:18	
Air Blank		0.000	06:18	
Control Test		0.080	06:19	
Air Blank		0.000	06:19	
Diagnostics Check		OK	06:19	

Data Download Date/Time: 06/27/2017 14:38

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: STANLEY,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:13

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 06/18/2017 20:43:50

Diagnostics Check		OK	20:34
Air Blank		0.000	20:34
Control Test		0.079	20:34
Air Blank		0.000	20:35
Subject Sample	# 1	0.290	20:36
Breath Volume		2.925	20:36
Air Blank		0.000	20:36
Air Blank		0.000	20:38
Subject Sample	# 2	0.261	20:39
Breath Volume		1.839	20:39
Air Blank		0.000	20:39
Air Blank		0.000	20:41
Subject Sample	# 3	0.265	20:42
Breath Volume		2.527	20:42
Air Blank		0.000	20:42
Control Test		0.079	20:43
Air Blank		0.000	20:43
Diagnostics Check		OK	20:43

Data Download Date/Time: 06/27/2017 14:38

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHITE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:17

SUBJECT'S LAST NAME: MATHIS

DATE/TIME OF TEST: 06/19/2017 01:40:39

Diagnostics Check		OK	01:37	
Air Blank		0.000	01:38	
Control Test		0.079	01:38	
Air Blank		0.000	01:39	
Subject Sample	# 1		01:39	Subject Test Refused
Breath Volume		0.000	01:39	
Air Blank		0.000	01:39	
Control Test		0.079	01:40	
Air Blank		0.000	01:40	
Diagnostics Check		OK	01:40	

Data Download Date/Time: 06/27/2017 14:38

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HOOVER,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:51

SUBJECT'S LAST NAME: LAMPKIN

DATE/TIME OF TEST: 06/19/2017 02:17:14

Diagnostics Check		OK	02:11
Air Blank		0.000	02:11
Control Test		0.079	02:12
Air Blank		0.000	02:12
Subject Sample	# 1	0.000	02:13
Breath Volume		1.792	02:13
Air Blank		0.000	02:13
Air Blank		0.000	02:15
Subject Sample	# 2	0.000	02:15
Breath Volume		3.152	02:15
Air Blank		0.000	02:16
Control Test		0.079	02:16
Air Blank		0.000	02:17
Diagnostics Check		OK	02:17

Data Download Date/Time: 06/27/2017 14:38

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CAMACHO,K
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 06/15/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:02

SUBJECT'S LAST NAME: MCKEOWN

DATE/TIME OF TEST: 06/19/2017 04:28:53

Diagnostics Check		OK	04:22
Air Blank		0.000	04:23
Control Test		0.080	04:23
Air Blank		0.000	04:24
Subject Sample	# 1	0.153	04:24
Breath Volume		2.082	04:24
Air Blank		0.000	04:25
Air Blank		0.000	04:27
Subject Sample	# 2	0.155	04:27
Breath Volume		2.636	04:27
Air Blank		0.000	04:27
Control Test		0.080	04:28
Air Blank		0.000	04:28
Diagnostics Check		OK	04:28

Data Download Date/Time: 06/27/2017 14:38

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MOSLEY,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:57

SUBJECT'S LAST NAME: ALMEIDA

DATE/TIME OF TEST: 06/20/2017 01:23:57

Diagnostics Check		OK	01:17
Air Blank		0.000	01:18
Control Test		0.079	01:18
Air Blank		0.000	01:19
Subject Sample	# 1	0.000	01:19
Breath Volume		1.523	01:19
Air Blank		0.000	01:20
Air Blank		0.000	01:22
Subject Sample	# 2	0.000	01:22
Breath Volume		2.812	01:22
Air Blank		0.000	01:23
Control Test		0.079	01:23
Air Blank		0.000	01:23
Diagnostics Check		OK	01:23

Data Download Date/Time: 06/27/2017 14:38

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MOELLER,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:57

SUBJECT'S LAST NAME: PACE

DATE/TIME OF TEST: 06/20/2017 03:39:07

Diagnostics Check		OK	03:36	
Air Blank		0.000	03:36	
Control Test		0.080	03:36	
Air Blank		0.000	03:37	
Subject Sample	# 1		03:37	Subject Test Refused
Breath Volume		0.000	03:37	
Air Blank		0.000	03:38	
Control Test		0.080	03:38	
Air Blank		0.000	03:39	
Diagnostics Check		OK	03:39	

Data Download Date/Time: 06/27/2017 14:38

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BRYANT,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: MARSHALL

DATE/TIME OF TEST: 06/21/2017 04:06:50

Diagnostics Check		OK	04:02	
Air Blank		0.000	04:02	
Control Test		0.080	04:03	
Air Blank		0.000	04:03	
Subject Sample	# 1		04:05	Subject Test Refused
Breath Volume		0.000	04:05	
Air Blank		0.000	04:05	
Control Test		0.079	04:06	
Air Blank		0.000	04:06	
Diagnostics Check		OK	04:06	

Data Download Date/Time: 06/27/2017 14:38

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MARRERO,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:26

SUBJECT'S LAST NAME: MEADOWS

DATE/TIME OF TEST: 06/21/2017 21:49:07

Diagnostics Check		OK	21:46	
Air Blank		0.000	21:46	
Control Test		0.080	21:47	
Air Blank		0.000	21:47	
Subject Sample	# 1		21:47	Subject Test Refused
Breath Volume		0.000	21:47	
Air Blank		0.000	21:48	
Control Test		0.080	21:48	
Air Blank		0.000	21:49	
Diagnostics Check		OK	21:49	

Data Download Date/Time: 06/27/2017 14:38

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: STANLEY,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: BENTON

DATE/TIME OF TEST: 06/28/2017 00:16:36

Diagnostics Check		OK	00:10
Air Blank		0.000	00:10
Control Test		0.079	00:11
Air Blank		0.000	00:11
Subject Sample	# 1	0.212	00:12
Breath Volume		3.136	00:12
Air Blank		0.000	00:12
Air Blank		0.000	00:14
Subject Sample	# 2	0.213	00:15
Breath Volume		3.804	00:15
Air Blank		0.000	00:15
Control Test		0.080	00:16
Air Blank		0.000	00:16
Diagnostics Check		OK	00:16

Data Download Date/Time: 07/13/2017 14:56

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHITE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: SWANN

DATE/TIME OF TEST: 06/28/2017 02:33:35

Diagnostics Check		OK	02:30	
Air Blank		0.000	02:31	
Control Test		0.080	02:31	
Air Blank		0.000	02:32	
Subject Sample	# 1		02:32	Subject Test Refused
Breath Volume		0.000	02:32	
Air Blank		0.000	02:32	
Control Test		0.080	02:33	
Air Blank		0.000	02:33	
Diagnostics Check		OK	02:33	

Data Download Date/Time: 07/13/2017 14:56

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MOELLER,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: MILLETE

DATE/TIME OF TEST: 06/29/2017 03:07:59

Diagnostics Check		OK	03:05	
Air Blank		0.000	03:05	
Control Test		0.080	03:05	
Air Blank		0.000	03:06	
Subject Sample	# 1		03:06	Subject Test Refused
Breath Volume		0.000	03:06	
Air Blank		0.000	03:07	
Control Test		0.079	03:07	
Air Blank		0.000	03:07	
Diagnostics Check		OK	03:07	

Data Download Date/Time: 07/13/2017 14:56

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FARLEY,J
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 06/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: WEILERSBACHER

DATE/TIME OF TEST: 06/30/2017 02:21:56

Diagnostics Check		OK	02:16
Air Blank		0.000	02:16
Control Test		0.080	02:16
Air Blank		0.000	02:17
Subject Sample	# 1	0.000	02:17
Breath Volume		2.367	02:17
Air Blank		0.000	02:18
Air Blank		0.000	02:20
Subject Sample	# 2	0.000	02:20
Breath Volume		3.062	02:20
Air Blank		0.000	02:21
Control Test		0.080	02:21
Air Blank		0.000	02:21
Diagnostics Check		OK	02:21

Data Download Date/Time: 07/13/2017 14:56

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHITE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:00

SUBJECT'S LAST NAME: COLLIER

DATE/TIME OF TEST: 06/19/2017 05:26:26

Diagnostics Check		OK	05:20
Air Blank		0.000	05:20
Control Test		0.078	05:20
Air Blank		0.000	05:21
Subject Sample	# 1	0.141	05:21
Breath Volume		1.406	05:21
Air Blank		0.000	05:22
Air Blank		0.000	05:24
Subject Sample	# 2	0.141	05:24
Breath Volume		1.957	05:24
Air Blank		0.000	05:25
Control Test		0.079	05:25
Air Blank		0.000	05:26
Diagnostics Check		OK	05:26

Data Download Date/Time: 06/27/2017 14:57

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAYLOR,TIMOTHY,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHITE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 18:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: BARD

DATE/TIME OF TEST: 06/19/2017 20:27:32

Diagnostics Check		OK	20:21
Air Blank		0.000	20:21
Control Test		0.079	20:21
Air Blank		0.000	20:22
Subject Sample	# 1	0.265	20:22
Breath Volume		2.207	20:22
Air Blank		0.000	20:23
Air Blank		0.000	20:25
Subject Sample	# 2	0.266	20:26
Breath Volume		2.390	20:26
Air Blank		0.000	20:26
Control Test		0.078	20:26
Air Blank		0.000	20:27
Diagnostics Check		OK	20:27

Data Download Date/Time: 06/27/2017 14:58

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAYLOR,TIMOTHY,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MORRISON,M.A
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 06/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 06/21/2017 02:07:30

Diagnostics Check		OK	02:01
Air Blank		0.000	02:01
Control Test		0.079	02:01
Air Blank		0.000	02:02
Subject Sample	# 1	0.099	02:02
Breath Volume		4.070	02:02
Air Blank		0.000	02:03
Air Blank		0.000	02:05
Subject Sample	# 2	0.101	02:06
Breath Volume		4.089	02:06
Air Blank		0.000	02:06
Control Test		0.079	02:06
Air Blank		0.000	02:07
Diagnostics Check		OK	02:07

Data Download Date/Time: 06/27/2017 14:58

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAYLOR,TIMOTHY,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: TORRES,W.E
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 06/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: LOPES

DATE/TIME OF TEST: 06/21/2017 04:48:50

Diagnostics Check		OK	04:40
Air Blank		0.000	04:40
Control Test		0.079	04:41
Air Blank		0.000	04:41
Subject Sample	# 1	0.182	04:42
Breath Volume		2.007	04:42
Air Blank		0.000	04:42
Air Blank		0.000	04:44
Subject Sample	# 2	0.171	04:47
Breath Volume		1.132	04:47
Air Blank		0.000	04:47
Control Test		0.078	04:48
Air Blank		0.000	04:48
Diagnostics Check		OK	04:48

Data Download Date/Time: 06/27/2017 14:58

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MULLIS,M,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ARCHBOLD,N
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:23

SUBJECT'S LAST NAME: PRESCOTT

DATE/TIME OF TEST: 06/22/2017 05:48:59

Diagnostics Check		OK	05:46	
Air Blank		0.000	05:46	
Control Test		0.079	05:46	
Air Blank		0.000	05:47	
Subject Sample	# 1		05:47	Subject Test Refused
Breath Volume		0.000	05:47	
Air Blank		0.000	05:48	
Control Test		0.079	05:48	
Air Blank		0.000	05:48	
Diagnostics Check		OK	05:48	

Data Download Date/Time: 06/27/2017 14:58

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY,SHANON,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BAI,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:27

SUBJECT'S LAST NAME: REPYNECK

DATE/TIME OF TEST: 06/24/2017 00:54:23

Diagnostics Check		OK	00:48
Air Blank		0.000	00:48
Control Test		0.080	00:49
Air Blank		0.000	00:49
Subject Sample	# 1	0.195	00:49
Breath Volume		2.769	00:49
Air Blank		0.000	00:50
Air Blank		0.000	00:52
Subject Sample	# 2	0.192	00:52
Breath Volume		3.031	00:52
Air Blank		0.000	00:53
Control Test		0.079	00:53
Air Blank		0.000	00:54
Diagnostics Check		OK	00:54

Data Download Date/Time: 06/27/2017 14:58

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY,SHANON,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CAMACHO,K
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 06/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:02

SUBJECT'S LAST NAME: LAING

DATE/TIME OF TEST: 06/24/2017 03:29:49

Diagnostics Check		OK	03:23
Air Blank		0.000	03:23
Control Test		0.079	03:23
Air Blank		0.000	03:24
Subject Sample	# 1	0.106	03:25
Breath Volume		3.855	03:25
Air Blank		0.000	03:25
Air Blank		0.000	03:27
Subject Sample	# 2	0.105	03:28
Breath Volume		3.957	03:28
Air Blank		0.000	03:28
Control Test		0.078	03:29
Air Blank		0.000	03:29
Diagnostics Check		OK	03:29

Data Download Date/Time: 06/27/2017 14:58

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY,SHANON,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: GARCIA,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:28

SUBJECT'S LAST NAME: MARICIO

DATE/TIME OF TEST: 06/25/2017 02:55:37

Diagnostics Check		OK	02:48
Air Blank		0.000	02:49
Control Test		0.079	02:49
Air Blank		0.000	02:49
Subject Sample	# 1	0.171	02:50
Breath Volume		1.429	02:50
Air Blank		0.000	02:51
Air Blank		0.000	02:53
Subject Sample	# 2	0.180	02:54
Breath Volume		1.218	02:54
Air Blank		0.000	02:54
Control Test		0.079	02:55
Air Blank		0.000	02:55
Diagnostics Check		OK	02:55

Data Download Date/Time: 06/27/2017 14:58

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY,SHANON,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HESS,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:43

SUBJECT'S LAST NAME: MOLINA

DATE/TIME OF TEST: 06/25/2017 04:12:14

Diagnostics Check		OK	04:05
Air Blank		0.000	04:06
Control Test		0.079	04:06
Air Blank		0.000	04:07
Subject Sample	# 1	0.143	04:07
Breath Volume		2.425	04:07
Air Blank		0.000	04:08
Air Blank		0.000	04:10
Subject Sample	# 2	0.142	04:10
Breath Volume		3.003	04:10
Air Blank		0.000	04:11
Control Test		0.079	04:11
Air Blank		0.000	04:12
Diagnostics Check		OK	04:12

Data Download Date/Time: 06/27/2017 14:58

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY,SHANON,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MOELLER,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:19

SUBJECT'S LAST NAME: JANKOWSKI

DATE/TIME OF TEST: 06/26/2017 02:49:42

Diagnostics Check		OK	02:40
Air Blank		0.000	02:40
Control Test		0.079	02:41
Air Blank		0.000	02:41
Subject Sample	# 1	0.173	02:42
Breath Volume		4.062	02:42
Air Blank		0.000	02:42
Air Blank		0.000	02:44
Subject Sample	# 2	0.149	02:45
Breath Volume		1.257	02:45
Air Blank		0.000	02:45
Air Blank		0.000	02:47
Subject Sample	# 3	0.161	02:48
Breath Volume		2.960	02:48
Air Blank		0.000	02:48
Control Test		0.078	02:49
Air Blank		0.000	02:49
Diagnostics Check		OK	02:49

Data Download Date/Time: 06/27/2017 14:58

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MURRIELL,E,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: STANLEY,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 17:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:30

SUBJECT'S LAST NAME: BOONE

DATE/TIME OF TEST: 06/26/2017 18:54:01

Diagnostics Check		OK	18:51	
Air Blank		0.000	18:51	
Control Test		0.078	18:51	
Air Blank		0.000	18:52	
Subject Sample	# 1		18:52	Subject Test Refused
Breath Volume		0.000	18:52	
Air Blank		0.000	18:53	
Control Test		0.079	18:53	
Air Blank		0.000	18:53	
Diagnostics Check		OK	18:54	

Data Download Date/Time: 06/27/2017 14:58

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY,SHANON,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CAMACHO,K
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 06/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:33

SUBJECT'S LAST NAME: LUGO

DATE/TIME OF TEST: 06/27/2017 01:00:05

Diagnostics Check		OK	00:53
Air Blank		0.000	00:53
Control Test		0.079	00:54
Air Blank		0.000	00:54
Subject Sample	# 1	0.143	00:55
Breath Volume		2.476	00:55
Air Blank		0.000	00:56
Air Blank		0.000	00:58
Subject Sample	# 2	0.154	00:58
Breath Volume		2.250	00:58
Air Blank		0.000	00:59
Control Test		0.080	00:59
Air Blank		0.000	01:00
Diagnostics Check		OK	01:00

Data Download Date/Time: 06/27/2017 15:00

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY,SHANON,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BRETT,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:09

SUBJECT'S LAST NAME: CLARK

DATE/TIME OF TEST: 06/27/2017 03:36:03

Diagnostics Check		OK	03:29
Air Blank		0.000	03:30
Control Test		0.079	03:30
Air Blank		0.000	03:30
Subject Sample	# 1	0.000	03:31
Breath Volume		3.656	03:31
Air Blank		0.000	03:31
Air Blank		0.000	03:33
Subject Sample	# 2	0.000	03:34
Breath Volume		2.269	03:34
Air Blank		0.000	03:35
Control Test		0.079	03:35
Air Blank		0.000	03:35
Diagnostics Check		OK	03:36

Data Download Date/Time: 06/27/2017 15:00

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAUSEY,JOHN,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ANDRES,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 15:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:41

SUBJECT'S LAST NAME: LUTHER

DATE/TIME OF TEST: 06/27/2017 18:04:41

Diagnostics Check		OK	18:01	
Air Blank		0.000	18:02	
Control Test		0.077	18:02	
Air Blank		0.000	18:02	
Subject Sample	# 1		18:03	Subject Test Refused
Breath Volume		0.000	18:03	
Air Blank		0.000	18:03	
Control Test		0.078	18:04	
Air Blank		0.000	18:04	
Diagnostics Check		OK	18:04	

Data Download Date/Time: 07/13/2017 15:15

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAYLOR,TIMOTHY,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WOOLDRIDGE,D.C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 14:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:50

SUBJECT'S LAST NAME: BAKER

DATE/TIME OF TEST: 06/30/2017 16:17:19

Diagnostics Check		OK	16:10
Air Blank		0.000	16:10
Control Test		0.079	16:11
Air Blank		0.000	16:11
Subject Sample	# 1	0.247	16:12
Breath Volume		2.996	16:12
Air Blank		0.000	16:13
Air Blank		0.000	16:15
Subject Sample	# 2	0.244	16:15
Breath Volume		2.398	16:15
Air Blank		0.000	16:16
Control Test		0.078	16:16
Air Blank		0.000	16:17
Diagnostics Check		OK	16:17

Data Download Date/Time: 07/13/2017 15:15

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOSTON,D,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: EANES,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:29

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 06/03/2017 23:58:56

Diagnostics Check		OK	23:52
Air Blank		0.000	23:52
Control Test		0.078	23:53
Air Blank		0.000	23:53
Subject Sample	# 1	0.176	23:54
Breath Volume		1.199	23:54
Air Blank		0.000	23:54
Air Blank		0.000	23:56
Subject Sample	# 2	0.169	23:57
Breath Volume		1.394	23:57
Air Blank		0.000	23:58
Control Test		0.078	23:58
Air Blank		0.000	23:58
Diagnostics Check		OK	23:58

Data Download Date/Time: 06/15/2017 12:58

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOSTON,D,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BYRNE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:03

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 06/05/2017 00:31:06

Diagnostics Check		OK	00:23
Air Blank		0.000	00:24
Control Test		0.079	00:24
Air Blank		0.000	00:25
Subject Sample	# 1	0.213	00:26
Breath Volume		2.242	00:26
Air Blank		0.000	00:27
Air Blank		0.000	00:28
Subject Sample	# 2	0.224	00:29
Breath Volume		2.468	00:29
Air Blank		0.000	00:30
Control Test		0.078	00:30
Air Blank		0.000	00:31
Diagnostics Check		OK	00:31

Data Download Date/Time: 06/15/2017 12:58

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOSTON,D,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MARRERO,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: PUPO SIERRA

DATE/TIME OF TEST: 06/05/2017 01:43:12

Diagnostics Check		OK	01:35
Air Blank		0.000	01:36
Control Test		0.079	01:36
Air Blank		0.000	01:36
Subject Sample	# 1	0.114	01:38
Breath Volume		1.726	01:38
Air Blank		0.000	01:38
Air Blank		0.000	01:40
Subject Sample	# 2	0.110	01:41
Breath Volume		1.410	01:41
Air Blank		0.000	01:42
Control Test		0.079	01:42
Air Blank		0.000	01:43
Diagnostics Check		OK	01:43

Data Download Date/Time: 06/15/2017 12:58

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOSTON,D,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BROWN,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:21

SUBJECT'S LAST NAME: FLORES

DATE/TIME OF TEST: 06/06/2017 01:49:12

Diagnostics Check		OK	01:42
Air Blank		0.000	01:42
Control Test		0.078	01:42
Air Blank		0.000	01:43
Subject Sample	# 1	0.122	01:44
Breath Volume		1.886	01:44
Air Blank		0.000	01:45
Air Blank		0.000	01:47
Subject Sample	# 2	0.123	01:47
Breath Volume		1.671	01:47
Air Blank		0.000	01:48
Control Test		0.078	01:48
Air Blank		0.000	01:49
Diagnostics Check		OK	01:49

Data Download Date/Time: 06/15/2017 12:58

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOSTON,D,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MARRERO,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: CLOUD

DATE/TIME OF TEST: 06/14/2017 00:13:28

Diagnostics Check		OK	00:05
Air Blank		0.000	00:05
Control Test		0.079	00:06
Air Blank		0.000	00:06
Subject Sample	# 1	0.075	00:08
Breath Volume		1.753	00:08
Air Blank		0.000	00:08
Air Blank		0.000	00:10
Subject Sample	# 2	0.076	00:11
Breath Volume		1.898	00:11
Air Blank		0.000	00:12
Control Test		0.079	00:12
Air Blank		0.000	00:13
Diagnostics Check		OK	00:13

Data Download Date/Time: 06/15/2017 12:58

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOSTON,D,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BRYANT,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:33

SUBJECT'S LAST NAME: DOMINGO

DATE/TIME OF TEST: 06/14/2017 04:03:32

Diagnostics Check		OK	03:54
Air Blank		0.000	03:55
Control Test		0.078	03:55
Air Blank		0.000	03:55
Subject Sample	# 1	0.149	03:58
Breath Volume		2.550	03:58
Air Blank		0.000	03:58
Air Blank		0.000	04:00
Subject Sample	# 2	0.152	04:01
Breath Volume		2.835	04:01
Air Blank		0.000	04:02
Control Test		0.077	04:02
Air Blank		0.000	04:03
Diagnostics Check		OK	04:03

Data Download Date/Time: 06/15/2017 12:58

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MULLIS,M,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHITE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: SMOTHERS

DATE/TIME OF TEST: 06/18/2017 03:53:03

Diagnostics Check		OK	03:46
Air Blank		0.000	03:46
Control Test		0.079	03:46
Air Blank		0.000	03:47
Subject Sample	# 1	0.086	03:48
Breath Volume		3.234	03:48
Air Blank		0.000	03:49
Air Blank		0.000	03:50
Subject Sample	# 2	0.078	03:51
Breath Volume		1.914	03:51
Air Blank		0.000	03:52
Control Test		0.079	03:52
Air Blank		0.000	03:52
Diagnostics Check		OK	03:53

Data Download Date/Time: 06/27/2017 14:54

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MULLIS,M,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ZURN,E
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 05:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:18

SUBJECT'S LAST NAME: MOSHER

DATE/TIME OF TEST: 06/18/2017 05:44:48

Diagnostics Check		OK	05:38
Air Blank		0.000	05:38
Control Test		0.079	05:39
Air Blank		0.000	05:39
Subject Sample	# 1	0.165	05:40
Breath Volume		3.628	05:40
Air Blank		0.000	05:40
Air Blank		0.000	05:42
Subject Sample	# 2	0.164	05:43
Breath Volume		3.753	05:43
Air Blank		0.000	05:43
Control Test		0.078	05:44
Air Blank		0.000	05:44
Diagnostics Check		OK	05:44

Data Download Date/Time: 06/27/2017 14:54

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MULLIS,M,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FARLEY,J
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 06/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:18

SUBJECT'S LAST NAME: SEYMOUR

DATE/TIME OF TEST: 06/22/2017 02:44:39

Diagnostics Check		OK	02:38
Air Blank		0.000	02:38
Control Test		0.078	02:39
Air Blank		0.000	02:39
Subject Sample	# 1	0.101	02:40
Breath Volume		2.195	02:40
Air Blank		0.000	02:40
Air Blank		0.000	02:42
Subject Sample	# 2	0.099	02:43
Breath Volume		2.714	02:43
Air Blank		0.000	02:43
Control Test		0.079	02:44
Air Blank		0.000	02:44
Diagnostics Check		OK	02:44

Data Download Date/Time: 06/27/2017 14:54

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MULLIS,M,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FARLEY,J
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 06/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:42

SUBJECT'S LAST NAME: MORRIS

DATE/TIME OF TEST: 06/22/2017 05:06:49

Diagnostics Check		OK	05:03	
Air Blank		0.000	05:04	
Control Test		0.079	05:04	
Air Blank		0.000	05:05	
Subject Sample	# 1		05:05	Subject Test Refused
Breath Volume		0.000	05:05	
Air Blank		0.000	05:05	
Control Test		0.079	05:06	
Air Blank		0.000	05:06	
Diagnostics Check		OK	05:06	

Data Download Date/Time: 06/27/2017 14:54

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOSTON,D,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BYRNE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:13

SUBJECT'S LAST NAME: DEGUZMAN

DATE/TIME OF TEST: 06/23/2017 00:42:06

Diagnostics Check		OK	00:34
Air Blank		0.000	00:34
Control Test		0.080	00:34
Air Blank		0.000	00:35
Subject Sample	# 1	0.141	00:36
Breath Volume		4.597	00:36
Air Blank		0.000	00:37
Air Blank		0.000	00:38
Subject Sample	# 2	0.134	00:40
Breath Volume		2.933	00:40
Air Blank		0.000	00:41
Control Test		0.079	00:41
Air Blank		0.000	00:42
Diagnostics Check		OK	00:42

Data Download Date/Time: 06/27/2017 14:54

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOSTON,D,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FARLEY,J
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 06/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:59

SUBJECT'S LAST NAME: SOLANO

DATE/TIME OF TEST: 06/23/2017 04:29:12

Diagnostics Check		OK	04:20
Air Blank		0.000	04:20
Control Test		0.079	04:21
Air Blank		0.000	04:21
Subject Sample	# 1	0.000	04:22
Breath Volume		2.796	04:22
Air Blank		0.000	04:23
Air Blank		0.000	04:25
Subject Sample	# 2	0.000	04:27
Breath Volume		2.410	04:27
Air Blank		0.000	04:28
Control Test		0.079	04:28
Air Blank		0.000	04:29
Diagnostics Check		OK	04:29

Data Download Date/Time: 06/27/2017 14:54

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOSTON,D,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HOWELL,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:51

SUBJECT'S LAST NAME: RASOR

DATE/TIME OF TEST: 06/23/2017 06:34:10

Diagnostics Check		OK	06:30	
Air Blank		0.000	06:31	
Control Test		0.079	06:31	
Air Blank		0.000	06:32	
Subject Sample	# 1		06:32	Subject Test Refused
Breath Volume		0.000	06:32	
Air Blank		0.000	06:33	
Control Test		0.079	06:33	
Air Blank		0.000	06:34	
Diagnostics Check		OK	06:34	

Data Download Date/Time: 06/27/2017 14:54

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOSTON,D,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHITE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:41

SUBJECT'S LAST NAME: ALLEN

DATE/TIME OF TEST: 06/24/2017 00:09:38

Diagnostics Check		OK	00:01
Air Blank		0.000	00:01
Control Test		0.079	00:02
Air Blank		0.000	00:02
Subject Sample	# 1	0.119	00:04
Breath Volume		1.683	00:04
Air Blank		0.000	00:05
Air Blank		0.000	00:06
Subject Sample	# 2	0.117	00:08
Breath Volume		1.476	00:08
Air Blank		0.000	00:08
Control Test		0.078	00:09
Air Blank		0.000	00:09
Diagnostics Check		OK	00:09

Data Download Date/Time: 06/27/2017 14:54

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOSTON,D,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MARRERO,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:33

SUBJECT'S LAST NAME: BRYANT

DATE/TIME OF TEST: 06/24/2017 03:02:11

Diagnostics Check		OK	02:54
Air Blank		0.000	02:54
Control Test		0.079	02:55
Air Blank		0.000	02:55
Subject Sample	# 1	0.086	02:56
Breath Volume		1.476	02:56
Air Blank		0.000	02:57
Air Blank		0.000	02:59
Subject Sample	# 2	0.098	03:00
Breath Volume		2.511	03:00
Air Blank		0.000	03:01
Control Test		0.082	03:01
Air Blank		0.000	03:02
Diagnostics Check		OK	03:02

Data Download Date/Time: 06/27/2017 14:54

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOSTON,D,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHITE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:37

SUBJECT'S LAST NAME: CROWDER

DATE/TIME OF TEST: 06/25/2017 03:05:03

Diagnostics Check		OK	02:57
Air Blank		0.000	02:58
Control Test		0.080	02:58
Air Blank		0.000	02:58
Subject Sample	# 1	0.148	03:00
Breath Volume		2.054	03:00
Air Blank		0.000	03:00
Air Blank		0.000	03:02
Subject Sample	# 2	0.148	03:03
Breath Volume		2.750	03:03
Air Blank		0.000	03:04
Control Test		0.079	03:04
Air Blank		0.000	03:04
Diagnostics Check		OK	03:05

Data Download Date/Time: 06/27/2017 14:54

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOSTON,D,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HILLIER,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:07

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 06/25/2017 04:42:32

Diagnostics Check		OK	04:39	
Air Blank		0.000	04:40	
Control Test		0.080	04:40	
Air Blank		0.000	04:40	
Subject Sample	# 1		04:41	Subject Test Refused
Breath Volume		0.000	04:41	
Air Blank		0.000	04:41	
Control Test		0.080	04:41	
Air Blank		0.000	04:42	
Diagnostics Check		OK	04:42	

Data Download Date/Time: 06/27/2017 14:54

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOSTON,D,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHITE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:32

SUBJECT'S LAST NAME: MCINTIRE

DATE/TIME OF TEST: 06/26/2017 01:08:44

Diagnostics Check		OK	01:05	
Air Blank		0.000	01:06	
Control Test		0.079	01:06	
Air Blank		0.000	01:07	
Subject Sample	# 1		01:07	Subject Test Refused
Breath Volume		0.000	01:07	
Air Blank		0.000	01:07	
Control Test		0.079	01:08	
Air Blank		0.000	01:08	
Diagnostics Check		OK	01:08	

Data Download Date/Time: 06/27/2017 14:54

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOSTON,D,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MORRISON,M
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 06/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:06

SUBJECT'S LAST NAME: MCDILL

DATE/TIME OF TEST: 06/26/2017 05:57:52

Diagnostics Check		OK	05:55	
Air Blank		0.000	05:55	
Control Test		0.079	05:55	
Air Blank		0.000	05:56	
Subject Sample	# 1		05:56	Subject Test Refused
Breath Volume		0.000	05:56	
Air Blank		0.000	05:56	
Control Test		0.080	05:57	
Air Blank		0.000	05:57	
Diagnostics Check		OK	05:57	

Data Download Date/Time: 06/27/2017 14:54

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOSTON,D,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: OLIVER,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:57

SUBJECT'S LAST NAME: KLINE

DATE/TIME OF TEST: 06/27/2017 03:24:53

Diagnostics Check		OK	03:18
Air Blank		0.000	03:18
Control Test		0.079	03:19
Air Blank		0.000	03:19
Subject Sample	# 1	0.000	03:20
Breath Volume		2.308	03:20
Air Blank		0.000	03:20
Air Blank		0.000	03:22
Subject Sample	# 2	0.000	03:23
Breath Volume		1.980	03:23
Air Blank		0.000	03:23
Control Test		0.079	03:24
Air Blank		0.000	03:24
Diagnostics Check		OK	03:24

Data Download Date/Time: 06/27/2017 14:54

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GILL,T,V
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: MASON,C
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 05/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:14

SUBJECT'S LAST NAME: JOHNS

DATE/TIME OF TEST: 06/16/2017 00:57:12

Diagnostics Check		OK	00:50
Air Blank		0.000	00:50
Control Test		0.081	00:51
Air Blank		0.000	00:51
Subject Sample	# 1	0.211	00:52
Breath Volume		2.339	00:52
Air Blank		0.000	00:53
Air Blank		0.000	00:55
Subject Sample	# 2	0.217	00:55
Breath Volume		2.000	00:55
Air Blank		0.000	00:56
Control Test		0.080	00:56
Air Blank		0.000	00:57
Diagnostics Check		OK	00:57

Data Download Date/Time: 06/26/2017 08:27

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001282

INSTRUMENT REGISTERED TO: UNIVERSITY OF FL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAIN II,KENNETH,A
BREATH TEST OPERATOR AGENCY: U of Fl PD
ARRESTING OFFICER: BAIN II,KENNETH
ARRESTING OFFICER AGENCY: U of Fl PD
LAST AGENCY INSPECTION: 05/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A2
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:57

SUBJECT'S LAST NAME: HENRY

DATE/TIME OF TEST: 06/03/2017 03:28:24

Diagnostics Check		OK	03:22
Air Blank		0.000	03:22
Control Test		0.078	03:23
Air Blank		0.000	03:23
Subject Sample	# 1	0.107	03:23
Breath Volume		2.195	03:23
Air Blank		0.000	03:24
Air Blank		0.000	03:26
Subject Sample	# 2	0.106	03:26
Breath Volume		1.925	03:26
Air Blank		0.000	03:27
Control Test		0.079	03:27
Air Blank		0.000	03:28
Diagnostics Check		OK	03:28

Data Download Date/Time: 06/28/2017 10:03

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001282

INSTRUMENT REGISTERED TO: UNIVERSITY OF FL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLWAY,CHAD,S
BREATH TEST OPERATOR AGENCY: U of Fl PD
ARRESTING OFFICER: NARAYAN,ROB
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 05/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A2
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 06:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:50

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 06/11/2017 07:21:10

Diagnostics Check		OK	07:13
Air Blank		0.000	07:14
Control Test		0.078	07:14
Air Blank		0.000	07:14
Subject Sample	# 1	0.106	07:16
Breath Volume		1.238	07:16
Air Blank		0.000	07:16
Air Blank		0.000	07:18
Subject Sample	# 2	0.109	07:19
Breath Volume		1.683	07:19
Air Blank		0.000	07:20
Control Test		0.079	07:20
Air Blank		0.000	07:21
Diagnostics Check		OK	07:21

Data Download Date/Time: 06/28/2017 10:03

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001282

INSTRUMENT REGISTERED TO: UNIVERSITY OF FL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEWART,DILLON,J
BREATH TEST OPERATOR AGENCY: U of Fl PD
ARRESTING OFFICER: PASCALE,KAITLYN
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 05/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A2
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 18:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:06

SUBJECT'S LAST NAME: NICKLAS

DATE/TIME OF TEST: 06/20/2017 19:40:07

Diagnostics Check		OK	19:29	
Air Blank		0.000	19:29	
Control Test		0.077	19:30	
Air Blank		0.000	19:30	
Subject Sample	# 1	0.000	19:33	Volume Not Met
Breath Volume		0.054	19:33	
Air Blank		0.000	19:34	
Air Blank		0.000	19:36	
Subject Sample	# 2		19:38	Subject Test Refused
Breath Volume		0.105	19:38	
Air Blank		0.000	19:39	
Control Test		0.077	19:39	
Air Blank		0.000	19:40	
Diagnostics Check		OK	19:40	

Data Download Date/Time: 06/28/2017 10:03

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001282

INSTRUMENT REGISTERED TO: UNIVERSITY OF FL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZARATE,JESSICA,L
BREATH TEST OPERATOR AGENCY: U of Fl PD
ARRESTING OFFICER: SHEFFIELD,KHRISTOPHER
ARRESTING OFFICER AGENCY: U of Fl PD
LAST AGENCY INSPECTION: 05/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A2
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:20

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 06/24/2017 05:55:29

Diagnostics Check		OK	05:48
Air Blank		0.000	05:49
Control Test		0.079	05:49
Air Blank		0.000	05:49
Subject Sample	# 1	0.162	05:50
Breath Volume		3.308	05:50
Air Blank		0.000	05:51
Air Blank		0.000	05:53
Subject Sample	# 2	0.161	05:53
Breath Volume		3.601	05:53
Air Blank		0.000	05:54
Control Test		0.078	05:54
Air Blank		0.000	05:55
Diagnostics Check		OK	05:55

Data Download Date/Time: 06/28/2017 10:03

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001282

INSTRUMENT REGISTERED TO: UNIVERSITY OF FL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEWART,DILLON,J
BREATH TEST OPERATOR AGENCY: U of Fl PD
ARRESTING OFFICER: SHEFFIELD,KHRISTOPHER
ARRESTING OFFICER AGENCY: U of Fl PD
LAST AGENCY INSPECTION: 05/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A2
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:51

SUBJECT'S LAST NAME: FORSYTH

DATE/TIME OF TEST: 06/26/2017 02:23:52

Diagnostics Check		OK	02:17
Air Blank		0.000	02:17
Control Test		0.078	02:18
Air Blank		0.000	02:18
Subject Sample	# 1	0.055	02:19
Breath Volume		1.335	02:19
Air Blank		0.000	02:20
Air Blank		0.000	02:21
Subject Sample	# 2	0.054	02:22
Breath Volume		1.128	02:22
Air Blank		0.000	02:22
Control Test		0.076	02:23
Air Blank		0.000	02:23
Diagnostics Check		OK	02:23

Data Download Date/Time: 06/28/2017 10:03

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001282

INSTRUMENT REGISTERED TO: UNIVERSITY OF FL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRICKER,JEFFREY,S
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: FRICKER,JEFFREY
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 06/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A2
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 07:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:25

SUBJECT'S LAST NAME: BEZET

DATE/TIME OF TEST: 06/30/2017 09:01:00

Diagnostics Check		OK	08:54
Air Blank		0.000	08:55
Control Test		0.078	08:55
Air Blank		0.000	08:56
Subject Sample	# 1	0.195	08:56
Breath Volume		1.753	08:56
Air Blank		0.000	08:57
Air Blank		0.000	08:59
Subject Sample	# 2	0.183	08:59
Breath Volume		1.339	08:59
Air Blank		0.000	09:00
Control Test		0.078	09:00
Air Blank		0.000	09:00
Diagnostics Check		OK	09:01

Data Download Date/Time: 07/27/2017 10:18

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001283

INSTRUMENT REGISTERED TO: HIGH SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRESNELL,ETHAN,C
BREATH TEST OPERATOR AGENCY: High Springs PD
ARRESTING OFFICER: PRESNELL,ETHAN
ARRESTING OFFICER AGENCY: High Springs PD
LAST AGENCY INSPECTION: 05/23/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:01

SUBJECT'S LAST NAME: ROBERTS

DATE/TIME OF TEST: 06/11/2017 02:24:53

Diagnostics Check		OK	02:21	
Air Blank		0.000	02:22	
Control Test		0.081	02:22	
Air Blank		0.000	02:23	
Subject Sample	# 1		02:23	Subject Test Refused
Breath Volume		0.000	02:23	
Air Blank		0.000	02:23	
Control Test		0.081	02:24	
Air Blank		0.000	02:24	
Diagnostics Check		OK	02:24	

Data Download Date/Time: 06/30/2017 17:50

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001283

INSTRUMENT REGISTERED TO: HIGH SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YAKUBSIN,JAMES,R
BREATH TEST OPERATOR AGENCY: High Springs PD
ARRESTING OFFICER: HAMPTON,KENDRICK
ARRESTING OFFICER AGENCY: High Springs PD
LAST AGENCY INSPECTION: 05/23/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 18:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:27

SUBJECT'S LAST NAME: STEPHENS

DATE/TIME OF TEST: 06/11/2017 18:56:30

Diagnostics Check		OK	18:47
Air Blank		0.000	18:48
Control Test		0.081	18:48
Air Blank		0.000	18:49
Subject Sample	# 1	0.221	18:49
Breath Volume		2.203	18:49
Air Blank		0.000	18:50
Air Blank		0.000	18:52
Subject Sample	# 2	0.206	18:54
Breath Volume		2.136	18:54
Air Blank		0.000	18:55
Control Test		0.078	18:55
Air Blank		0.000	18:56
Diagnostics Check		OK	18:56

Data Download Date/Time: 06/30/2017 17:50

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001285

INSTRUMENT REGISTERED TO: GILCHRIST COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THWEATT,CAMERON,J
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: THWEATT,CAMERON
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 05/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 259585
DRY GAS EXPIRY DATE: 01/12/2019
VIOLATION CODE: DUI
ARREST TIME: 19:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:36

SUBJECT'S LAST NAME: GLUHM

DATE/TIME OF TEST: 06/29/2017 20:03:01

Diagnostics Check		OK	19:56
Air Blank		0.000	19:57
Control Test		0.078	19:57
Air Blank		0.000	19:57
Subject Sample	# 1	0.192	19:58
Breath Volume		1.324	19:58
Air Blank		0.000	19:59
Air Blank		0.000	20:01
Subject Sample	# 2	0.193	20:01
Breath Volume		2.046	20:01
Air Blank		0.000	20:02
Control Test		0.078	20:02
Air Blank		0.000	20:02
Diagnostics Check		OK	20:03

Data Download Date/Time: 06/30/2017 15:54

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001287

INSTRUMENT REGISTERED TO: BAKER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRANNON,AARON,K
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: BRANNON,AARON
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 05/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:58

SUBJECT'S LAST NAME: VINSON

DATE/TIME OF TEST: 06/06/2017 01:29:23

Diagnostics Check		OK	01:22
Air Blank		0.000	01:23
Control Test		0.082	01:23
Air Blank		0.000	01:24
Subject Sample	# 1	0.153	01:24
Breath Volume		2.187	01:24
Air Blank		0.000	01:25
Air Blank		0.000	01:27
Subject Sample	# 2	0.148	01:27
Breath Volume		1.796	01:27
Air Blank		0.000	01:28
Control Test		0.081	01:28
Air Blank		0.000	01:29
Diagnostics Check		OK	01:29

Data Download Date/Time: 07/04/2017 01:36

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001287

INSTRUMENT REGISTERED TO: BAKER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARLEY,JOSEPH,E
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: FARLEY,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 05/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:17

SUBJECT'S LAST NAME: FOWLER

DATE/TIME OF TEST: 06/17/2017 22:45:03

Diagnostics Check		OK	22:38
Air Blank		0.000	22:39
Control Test		0.081	22:39
Air Blank		0.000	22:39
Subject Sample	# 1	0.187	22:40
Breath Volume		1.464	22:40
Air Blank		0.000	22:41
Air Blank		0.000	22:42
Subject Sample	# 2	0.196	22:43
Breath Volume		1.871	22:43
Air Blank		0.000	22:44
Control Test		0.081	22:44
Air Blank		0.000	22:44
Diagnostics Check		OK	22:45

Data Download Date/Time: 07/04/2017 01:36

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001287

INSTRUMENT REGISTERED TO: BAKER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARLEY,JOSEPH,E
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: FARLEY,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 05/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: HARVEY

DATE/TIME OF TEST: 06/18/2017 01:35:01

Diagnostics Check		OK	01:25
Air Blank		0.000	01:26
Control Test		0.080	01:26
Air Blank		0.000	01:27
Subject Sample	# 1	0.210	01:30
Breath Volume		1.644	01:30
Air Blank		0.000	01:30
Air Blank		0.000	01:32
Subject Sample	# 2	0.210	01:33
Breath Volume		1.292	01:33
Air Blank		0.000	01:34
Control Test		0.079	01:34
Air Blank		0.000	01:34
Diagnostics Check		OK	01:35

Data Download Date/Time: 07/04/2017 01:36

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001287

INSTRUMENT REGISTERED TO: BAKER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMATT,JUSTIN,M
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: SMATT,JUSTIN
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 05/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 12:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:19

SUBJECT'S LAST NAME: FRASER

DATE/TIME OF TEST: 06/29/2017 01:44:30

Diagnostics Check		OK	01:41	
Air Blank		0.000	01:42	
Control Test		0.082	01:42	
Air Blank		0.000	01:42	
Subject Sample	# 1		01:43	Subject Test Refused
Breath Volume		0.000	01:43	
Air Blank		0.000	01:43	
Control Test		0.082	01:43	
Air Blank		0.000	01:44	
Diagnostics Check		OK	01:44	

Data Download Date/Time: 07/04/2017 01:36

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001291

INSTRUMENT REGISTERED TO: BRADFORD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORRELL,JOSEPH,B
BREATH TEST OPERATOR AGENCY: Bradford County SO
ARRESTING OFFICER: MORRELL,JOSEPH
ARRESTING OFFICER AGENCY: Bradford County SO
LAST AGENCY INSPECTION: 05/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG619605
DRY GAS EXPIRY DATE: 07/14/2018
VIOLATION CODE: DUI
ARREST TIME: 03:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:46

SUBJECT'S LAST NAME: BARRON

DATE/TIME OF TEST: 06/24/2017 04:21:29

Diagnostics Check		OK	04:14
Air Blank		0.000	04:15
Control Test		0.079	04:15
Air Blank		0.000	04:16
Subject Sample	# 1	0.157	04:16
Breath Volume		1.363	04:16
Air Blank		0.000	04:17
Air Blank		0.000	04:19
Subject Sample	# 2	0.149	04:19
Breath Volume		1.464	04:19
Air Blank		0.000	04:20
Control Test		0.078	04:20
Air Blank		0.000	04:21
Diagnostics Check		OK	04:21

Data Download Date/Time: 06/24/2017 08:58

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001293

INSTRUMENT REGISTERED TO: WILLISTON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANDIS,BRYAN,S
BREATH TEST OPERATOR AGENCY: Williston PD
ARRESTING OFFICER: LANDIS,BRYAN
ARRESTING OFFICER AGENCY: Williston PD
LAST AGENCY INSPECTION: 05/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 03:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: DANIELS

DATE/TIME OF TEST: 06/02/2017 04:56:04

Diagnostics Check		OK	04:49
Air Blank		0.000	04:50
Control Test		0.080	04:50
Air Blank		0.000	04:50
Subject Sample	# 1	0.152	04:51
Breath Volume		1.128	04:51
Air Blank		0.000	04:52
Air Blank		0.000	04:53
Subject Sample	# 2	0.150	04:54
Breath Volume		3.753	04:54
Air Blank		0.000	04:55
Control Test		0.079	04:55
Air Blank		0.000	04:55
Diagnostics Check		OK	04:56

Data Download Date/Time: 07/06/2017 09:34

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001305

INSTRUMENT REGISTERED TO: MARIANNA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COX,DAVID,S
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: COX,DAVID
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 05/23/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 19:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:55

SUBJECT'S LAST NAME: CASMUS

DATE/TIME OF TEST: 06/05/2017 20:25:38

Diagnostics Check		OK	20:19
Air Blank		0.000	20:19
Control Test		0.081	20:20
Air Blank		0.000	20:20
Subject Sample	# 1	0.126	20:21
Breath Volume		4.429	20:21
Air Blank		0.000	20:21
Air Blank		0.000	20:23
Subject Sample	# 2	0.124	20:24
Breath Volume		4.160	20:24
Air Blank		0.000	20:24
Control Test		0.079	20:25
Air Blank		0.000	20:25
Diagnostics Check		OK	20:25

Data Download Date/Time: 06/29/2017 11:48

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001305

INSTRUMENT REGISTERED TO: MARIANNA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VASSALLO,STEVEN,A
BREATH TEST OPERATOR AGENCY: Marianna PD
ARRESTING OFFICER: ARROYO,BRANDON
ARRESTING OFFICER AGENCY: Marianna PD
LAST AGENCY INSPECTION: 05/23/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 23:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:42

SUBJECT'S LAST NAME: GRZEGORCZYK

DATE/TIME OF TEST: 06/09/2017 00:21:11

Diagnostics Check		OK	00:18	
Air Blank		0.000	00:18	
Control Test		0.081	00:19	
Air Blank		0.000	00:19	
Subject Sample	# 1		00:19	Subject Test Refused
Breath Volume		0.000	00:19	
Air Blank		0.000	00:20	
Control Test		0.080	00:20	
Air Blank		0.000	00:21	
Diagnostics Check		OK	00:21	

Data Download Date/Time: 06/29/2017 11:48

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001306

INSTRUMENT REGISTERED TO: MEXICO BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCSWINEY,RICHARD,A
BREATH TEST OPERATOR AGENCY: Mexico Beach PD
ARRESTING OFFICER: MCSWINEY,RICHARD
ARRESTING OFFICER AGENCY: Mexico Beach PD
LAST AGENCY INSPECTION: 05/03/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:32

SUBJECT'S LAST NAME: BOGGS

DATE/TIME OF TEST: 06/05/2017 21:10:43

Diagnostics Check		OK	21:04
Air Blank		0.000	21:04
Control Test		0.080	21:05
Air Blank		0.000	21:05
Subject Sample	# 1	0.055	21:06
Breath Volume		1.257	21:06
Air Blank		0.000	21:06
Air Blank		0.000	21:08
Subject Sample	# 2	0.057	21:09
Breath Volume		1.265	21:09
Air Blank		0.000	21:09
Control Test		0.081	21:10
Air Blank		0.000	21:10
Diagnostics Check		OK	21:10

Data Download Date/Time: 06/07/2017 00:49

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001306

INSTRUMENT REGISTERED TO: MEXICO BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAUVERSBURK,BRENNA,M
BREATH TEST OPERATOR AGENCY: Mexico Beach PD
ARRESTING OFFICER: HAUVERSBURK,BRENNA
ARRESTING OFFICER AGENCY: Mexico Beach PD
LAST AGENCY INSPECTION: 06/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 16:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:30

SUBJECT'S LAST NAME: WORD

DATE/TIME OF TEST: 06/24/2017 17:54:48

Diagnostics Check		OK	17:43	
Air Blank		0.000	17:43	
Control Test		0.079	17:44	
Air Blank		0.000	17:44	
Subject Sample	# 1		17:47	No Sample Provided
Breath Volume		0.000	17:47	
Air Blank		0.000	17:48	
Air Blank		0.000	17:50	
Subject Sample	# 2		17:53	No Sample Provided
Breath Volume		0.000	17:53	
Air Blank		0.000	17:53	
Control Test		0.079	17:54	
Air Blank		0.000	17:54	
Diagnostics Check		OK	17:54	

Data Download Date/Time: 07/07/2017 09:40

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CALASSO,ANTHONY,
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: KELLY,MATTHEW
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 05/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG628401
DRY GAS EXPIRY DATE: 10/10/2018
VIOLATION CODE: DUI
ARREST TIME: 12:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:45

SUBJECT'S LAST NAME: BUCK

DATE/TIME OF TEST: 06/02/2017 13:22:55

Diagnostics Check		OK	13:16
Air Blank		0.000	13:17
Control Test		0.079	13:17
Air Blank		0.000	13:18
Subject Sample	# 1	0.272	13:18
Breath Volume		1.324	13:18
Air Blank		0.000	13:19
Air Blank		0.000	13:20
Subject Sample	# 2	0.269	13:21
Breath Volume		1.886	13:21
Air Blank		0.000	13:22
Control Test		0.079	13:22
Air Blank		0.000	13:22
Diagnostics Check		OK	13:22

Data Download Date/Time: 06/04/2017 12:10

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BASFORD,NICOLE,M
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: BASFORD,NICOLE
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 05/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG628401
DRY GAS EXPIRY DATE: 10/10/2018
VIOLATION CODE: BUI
ARREST TIME: 18:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:30

SUBJECT'S LAST NAME: MELVILLE

DATE/TIME OF TEST: 06/04/2017 19:57:59

Diagnostics Check		OK	19:51
Air Blank		0.000	19:51
Control Test		0.079	19:52
Air Blank		0.000	19:52
Subject Sample	# 1	0.219	19:53
Breath Volume		2.531	19:53
Air Blank		0.000	19:53
Air Blank		0.000	19:55
Subject Sample	# 2	0.223	19:56
Breath Volume		2.285	19:56
Air Blank		0.000	19:57
Control Test		0.079	19:57
Air Blank		0.000	19:57
Diagnostics Check		OK	19:57

Data Download Date/Time: 06/29/2017 14:15

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALDERMAN,JOSEPH,L
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: ALDERMAN,JOSEPH
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 05/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG628401
DRY GAS EXPIRY DATE: 10/10/2018
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:51

SUBJECT'S LAST NAME: DIXON

DATE/TIME OF TEST: 06/05/2017 02:23:34

Diagnostics Check		OK	02:17
Air Blank		0.000	02:17
Control Test		0.081	02:17
Air Blank		0.000	02:18
Subject Sample	# 1	0.188	02:19
Breath Volume		2.207	02:19
Air Blank		0.000	02:19
Air Blank		0.000	02:21
Subject Sample	# 2	0.180	02:22
Breath Volume		2.222	02:22
Air Blank		0.000	02:22
Control Test		0.080	02:22
Air Blank		0.000	02:23
Diagnostics Check		OK	02:23

Data Download Date/Time: 06/29/2017 14:15

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WURST,JOSHUA,D
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: WURST,JOSHUA
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 05/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG628401
DRY GAS EXPIRY DATE: 10/10/2018
VIOLATION CODE: DUI
ARREST TIME: 19:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:40

SUBJECT'S LAST NAME: WEED

DATE/TIME OF TEST: 06/08/2017 21:16:11

Diagnostics Check		OK	21:09
Air Blank		0.000	21:10
Control Test		0.080	21:10
Air Blank		0.000	21:11
Subject Sample	# 1	0.303	21:11
Breath Volume		1.562	21:11
Air Blank		0.000	21:12
Air Blank		0.000	21:14
Subject Sample	# 2	0.314	21:14
Breath Volume		1.765	21:14
Air Blank		0.000	21:15
Control Test		0.080	21:15
Air Blank		0.000	21:16
Diagnostics Check		OK	21:16

Data Download Date/Time: 06/29/2017 14:15

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANCE,CHARLES,R
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: LANCE,CHARLES
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 05/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG628401
DRY GAS EXPIRY DATE: 10/10/2018
VIOLATION CODE: DUI
ARREST TIME: 22:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:33

SUBJECT'S LAST NAME: JOHNS

DATE/TIME OF TEST: 06/11/2017 00:00:31

Diagnostics Check		OK	23:54
Air Blank		0.000	23:54
Control Test		0.080	23:54
Air Blank		0.000	23:55
Subject Sample	# 1	0.135	23:56
Breath Volume		3.117	23:56
Air Blank		0.000	23:56
Air Blank		0.000	23:58
Subject Sample	# 2	0.133	23:59
Breath Volume		3.371	23:59
Air Blank		0.000	23:59
Control Test		0.079	23:59
Air Blank		0.000	00:00
Diagnostics Check		OK	00:00

Data Download Date/Time: 06/29/2017 14:15

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUINN,CAMERON,B
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: QUINN,CAMERON
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 05/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG628401
DRY GAS EXPIRY DATE: 10/10/2018
VIOLATION CODE: DUI
ARREST TIME: 00:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:24

SUBJECT'S LAST NAME: MCIVER

DATE/TIME OF TEST: 06/12/2017 01:51:03

Diagnostics Check		OK	01:44
Air Blank		0.000	01:44
Control Test		0.080	01:45
Air Blank		0.000	01:45
Subject Sample	# 1	0.053	01:46
Breath Volume		2.917	01:46
Air Blank		0.000	01:47
Air Blank		0.000	01:48
Subject Sample	# 2	0.053	01:49
Breath Volume		1.421	01:49
Air Blank		0.000	01:50
Control Test		0.080	01:50
Air Blank		0.000	01:50
Diagnostics Check		OK	01:51

Data Download Date/Time: 06/29/2017 14:15

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAGUADOG,SHAUN,J
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: MERRITT,DAKOTA
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 05/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG628401
DRY GAS EXPIRY DATE: 10/10/2018
VIOLATION CODE: DUI
ARREST TIME: 07:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:12

SUBJECT'S LAST NAME: PINO

DATE/TIME OF TEST: 06/14/2017 08:38:39

Diagnostics Check		OK	08:32
Air Blank		0.000	08:32
Control Test		0.080	08:33
Air Blank		0.000	08:33
Subject Sample	# 1	0.085	08:34
Breath Volume		1.156	08:34
Air Blank		0.000	08:34
Air Blank		0.000	08:36
Subject Sample	# 2	0.075	08:37
Breath Volume		1.214	08:37
Air Blank		0.000	08:37
Control Test		0.080	08:38
Air Blank		0.000	08:38
Diagnostics Check		OK	08:38

Data Download Date/Time: 06/29/2017 14:15

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRICE,CARLON,E
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: PRICE,CARLON
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 05/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG628401
DRY GAS EXPIRY DATE: 10/10/2018
VIOLATION CODE: DUI
ARREST TIME: 01:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:56

SUBJECT'S LAST NAME: COLLINS

DATE/TIME OF TEST: 06/21/2017 03:07:23

Diagnostics Check		OK	03:01
Air Blank		0.000	03:01
Control Test		0.080	03:02
Air Blank		0.000	03:02
Subject Sample	# 1	0.000	03:03
Breath Volume		2.480	03:03
Air Blank		0.000	03:03
Air Blank		0.000	03:05
Subject Sample	# 2	0.000	03:06
Breath Volume		2.605	03:06
Air Blank		0.000	03:06
Control Test		0.079	03:06
Air Blank		0.000	03:07
Diagnostics Check		OK	03:07

Data Download Date/Time: 06/29/2017 14:15

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUBEL,NEIL,F
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: SCHELL,GARY
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 05/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG628401
DRY GAS EXPIRY DATE: 10/10/2018
VIOLATION CODE: DUI
ARREST TIME: 22:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:18

SUBJECT'S LAST NAME: HOBBS

DATE/TIME OF TEST: 06/24/2017 00:44:54

Diagnostics Check		OK	00:38
Air Blank		0.000	00:39
Control Test		0.080	00:39
Air Blank		0.000	00:40
Subject Sample	# 1	0.173	00:40
Breath Volume		1.214	00:40
Air Blank		0.000	00:41
Air Blank		0.000	00:42
Subject Sample	# 2	0.170	00:43
Breath Volume		1.511	00:43
Air Blank		0.000	00:43
Control Test		0.080	00:44
Air Blank		0.000	00:44
Diagnostics Check		OK	00:44

Data Download Date/Time: 06/29/2017 14:15

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALDERMAN,JOSEPH,L
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: ALDERMAN,JOSEPH
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 05/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG628401
DRY GAS EXPIRY DATE: 10/10/2018
VIOLATION CODE: DUI
ARREST TIME: 22:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: BACHAMP

DATE/TIME OF TEST: 06/24/2017 23:12:11

Diagnostics Check		OK	23:05
Air Blank		0.000	23:06
Control Test		0.080	23:06
Air Blank		0.000	23:07
Subject Sample	# 1	0.309	23:07
Breath Volume		3.062	23:07
Air Blank		0.000	23:08
Air Blank		0.000	23:10
Subject Sample	# 2	0.298	23:10
Breath Volume		2.832	23:10
Air Blank		0.000	23:11
Control Test		0.079	23:11
Air Blank		0.000	23:12
Diagnostics Check		OK	23:12

Data Download Date/Time: 06/29/2017 14:15

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MURPHY,PATRICK,J
BREATH TEST OPERATOR AGENCY: FDLE
ARRESTING OFFICER: TEST,TEST
ARRESTING OFFICER AGENCY: FDLE
LAST AGENCY INSPECTION: 06/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG628401
DRY GAS EXPIRY DATE: 10/10/2018
VIOLATION CODE: SYS CHECK
ARREST TIME: 07:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:00

SUBJECT'S LAST NAME: TEST

DATE/TIME OF TEST: 06/19/2017 12:24:22

Diagnostics Check		OK	12:18
Air Blank		0.000	12:18
Control Test		0.078	12:19
Air Blank		0.000	12:19
Subject Sample	# 1	0.000	12:20
Breath Volume		1.539	12:20
Air Blank		0.000	12:20
Air Blank		0.000	12:22
Subject Sample	# 2	0.000	12:23
Breath Volume		2.406	12:23
Air Blank		0.000	12:23
Control Test		0.079	12:23
Air Blank		0.000	12:24
Diagnostics Check		OK	12:24

Data Download Date/Time: 06/29/2017 14:17

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001310

INSTRUMENT REGISTERED TO: NICEVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WRIGHT,GRANT,C
BREATH TEST OPERATOR AGENCY: Niceville PD
ARRESTING OFFICER: SULLIVAN,DANNY
ARRESTING OFFICER AGENCY: Niceville PD
LAST AGENCY INSPECTION: 05/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: DUI
ARREST TIME: 03:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: HANSEN

DATE/TIME OF TEST: 06/02/2017 04:26:41

Diagnostics Check		OK	04:19
Air Blank		0.000	04:19
Control Test		0.082	04:20
Air Blank		0.000	04:20
Subject Sample	# 1	0.203	04:21
Breath Volume		1.195	04:21
Air Blank		0.000	04:21
Air Blank		0.000	04:23
Subject Sample	# 2	0.209	04:25
Breath Volume		2.250	04:25
Air Blank		0.000	04:25
Control Test		0.084	04:26
Air Blank		0.000	04:26
Diagnostics Check		OK	04:26

Data Download Date/Time: 06/02/2017 08:59

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001310

INSTRUMENT REGISTERED TO: NICEVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SULLIVAN,DANNY,
BREATH TEST OPERATOR AGENCY: Niceville PD
ARRESTING OFFICER: SULLIVAN,DANNY
ARRESTING OFFICER AGENCY: Niceville PD
LAST AGENCY INSPECTION: 05/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: DUI
ARREST TIME: 04:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:39

SUBJECT'S LAST NAME: MENEZES

DATE/TIME OF TEST: 06/10/2017 06:30:05

Diagnostics Check		OK	06:27	
Air Blank		0.000	06:27	
Control Test		0.081	06:28	
Air Blank		0.000	06:28	
Subject Sample	# 1		06:28	Subject Test Refused
Breath Volume		0.000	06:28	
Air Blank		0.000	06:29	
Control Test		0.081	06:29	
Air Blank		0.000	06:30	
Diagnostics Check		OK	06:30	

Data Download Date/Time: 06/29/2017 05:52

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001310

INSTRUMENT REGISTERED TO: NICEVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEWIS,SHAWN,W
BREATH TEST OPERATOR AGENCY: Niceville PD
ARRESTING OFFICER: WRIGHT,GRANT
ARRESTING OFFICER AGENCY: Niceville PD
LAST AGENCY INSPECTION: 05/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: DUI
ARREST TIME: 02:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:03

SUBJECT'S LAST NAME: HINOJOSA

DATE/TIME OF TEST: 06/11/2017 03:28:11

Diagnostics Check		OK	03:25	
Air Blank		0.000	03:25	
Control Test		0.082	03:26	
Air Blank		0.000	03:26	
Subject Sample	# 1		03:26	Subject Test Refused
Breath Volume		0.000	03:26	
Air Blank		0.000	03:27	
Control Test		0.082	03:27	
Air Blank		0.000	03:28	
Diagnostics Check		OK	03:28	

Data Download Date/Time: 06/29/2017 05:52

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001310

INSTRUMENT REGISTERED TO: NICEVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARRICO,KELLY,D
BREATH TEST OPERATOR AGENCY: Niceville PD
ARRESTING OFFICER: CARRICO,KELLY
ARRESTING OFFICER AGENCY: Niceville PD
LAST AGENCY INSPECTION: 05/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: DUI
ARREST TIME: 18:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:55

SUBJECT'S LAST NAME: FULTON

DATE/TIME OF TEST: 06/14/2017 20:23:08

Diagnostics Check		OK	20:16
Air Blank		0.000	20:16
Control Test		0.081	20:16
Air Blank		0.000	20:17
Subject Sample	# 1	0.171	20:17
Breath Volume		1.636	20:17
Air Blank		0.000	20:18
Air Blank		0.000	20:20
Subject Sample	# 2	0.178	20:21
Breath Volume		1.460	20:21
Air Blank		0.000	20:22
Control Test		0.082	20:22
Air Blank		0.000	20:23
Diagnostics Check		OK	20:23

Data Download Date/Time: 06/29/2017 05:52

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001310

INSTRUMENT REGISTERED TO: NICEVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEWIS,SHAWN,W
BREATH TEST OPERATOR AGENCY: Niceville PD
ARRESTING OFFICER: SULLIVAN,DANNY
ARRESTING OFFICER AGENCY: Niceville PD
LAST AGENCY INSPECTION: 05/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: DUI
ARREST TIME: 02:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: LAMM

DATE/TIME OF TEST: 06/24/2017 03:10:10

Diagnostics Check		OK	03:07	
Air Blank		0.000	03:07	
Control Test		0.082	03:08	
Air Blank		0.000	03:08	
Subject Sample	# 1		03:08	Subject Test Refused
Breath Volume		0.000	03:08	
Air Blank		0.000	03:09	
Control Test		0.082	03:09	
Air Blank		0.000	03:10	
Diagnostics Check		OK	03:10	

Data Download Date/Time: 06/29/2017 05:53

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001310

INSTRUMENT REGISTERED TO: NICEVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEWIS,SHAWN,W
BREATH TEST OPERATOR AGENCY: Niceville PD
ARRESTING OFFICER: SULLIVAN,DANNY
ARRESTING OFFICER AGENCY: Niceville PD
LAST AGENCY INSPECTION: 05/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: DUI
ARREST TIME: 03:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:48

SUBJECT'S LAST NAME: PHELPS

DATE/TIME OF TEST: 06/25/2017 04:19:28

Diagnostics Check		OK	04:12
Air Blank		0.000	04:13
Control Test		0.081	04:13
Air Blank		0.000	04:14
Subject Sample	# 1	0.174	04:14
Breath Volume		2.242	04:14
Air Blank		0.000	04:15
Air Blank		0.000	04:17
Subject Sample	# 2	0.157	04:17
Breath Volume		1.195	04:17
Air Blank		0.000	04:18
Control Test		0.081	04:18
Air Blank		0.000	04:19
Diagnostics Check		OK	04:19

Data Download Date/Time: 06/29/2017 05:53

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANDERS,DAVID,O
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: SANDERS,DAVID
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 05/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163937
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 22:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:13

SUBJECT'S LAST NAME: OSTERHOFF

DATE/TIME OF TEST: 06/03/2017 23:40:27

Diagnostics Check		OK	23:36	
Air Blank		0.000	23:37	
Control Test		0.080	23:37	
Air Blank		0.000	23:38	
Subject Sample	# 1		23:39	Subject Test Refused
Breath Volume		0.000	23:39	
Air Blank		0.000	23:39	
Control Test		0.080	23:39	
Air Blank		0.000	23:40	
Diagnostics Check		OK	23:40	

Data Download Date/Time: 06/30/2017 11:59

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PATE,CHARLES,S
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: N/A,N/A
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 05/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163937
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: PROB/PAROL
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:15

SUBJECT'S LAST NAME: CADENHEAD

DATE/TIME OF TEST: 06/13/2017 13:51:09

Diagnostics Check		OK	13:45
Air Blank		0.000	13:45
Control Test		0.080	13:45
Air Blank		0.000	13:46
Subject Sample	# 1	0.000	13:46
Breath Volume		1.816	13:46
Air Blank		0.000	13:47
Air Blank		0.000	13:49
Subject Sample	# 2	0.000	13:49
Breath Volume		1.886	13:49
Air Blank		0.000	13:50
Control Test		0.081	13:50
Air Blank		0.000	13:51
Diagnostics Check		OK	13:51

Data Download Date/Time: 06/30/2017 11:59

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FUGATE,MITCHELL,W
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: FUGATE,MITCHELL
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 05/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 19:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:44

SUBJECT'S LAST NAME: KINGSLEY

DATE/TIME OF TEST: 06/24/2017 20:24:01

Diagnostics Check		OK	20:19	
Air Blank		0.000	20:19	
Control Test		0.080	20:19	
Air Blank		0.000	20:20	
Subject Sample	# 1		20:22	RFI Detect
Breath Volume		0.000	20:22	
Air Blank		0.000	20:23	
Control Test		0.080	20:23	
Air Blank		0.000	20:23	
Diagnostics Check		OK	20:24	

Data Download Date/Time: 06/30/2017 11:59

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FUGATE,MITCHELL,W
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: FUGATE,MITCHELL
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 05/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1634437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 19:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:44

SUBJECT'S LAST NAME: KINGSLEY

DATE/TIME OF TEST: 06/24/2017 20:33:45

Diagnostics Check		OK	20:26
Air Blank		0.000	20:27
Control Test		0.080	20:27
Air Blank		0.000	20:28
Subject Sample	# 1	0.218	20:29
Breath Volume		1.726	20:29
Air Blank		0.000	20:30
Air Blank		0.000	20:31
Subject Sample	# 2	0.217	20:32
Breath Volume		1.839	20:32
Air Blank		0.000	20:32
Control Test		0.077	20:33
Air Blank		0.000	20:33
Diagnostics Check		OK	20:33

Data Download Date/Time: 06/30/2017 11:59

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,JONATHAN,B
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: SMITH,JONATHAN
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 05/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: DUI
ARREST TIME: 20:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:13

SUBJECT'S LAST NAME: MITCHELL

DATE/TIME OF TEST: 06/01/2017 21:47:55

Diagnostics Check		OK	21:39	
Air Blank		0.000	21:40	
Control Test		0.077	21:40	
Air Blank		0.000	21:41	
Subject Sample	# 1	0.230	21:42	
Breath Volume		1.218	21:42	
Air Blank		0.000	21:42	
Air Blank		0.000	21:44	
Subject Sample	# 2	0.236	21:45	
Breath Volume		1.953	21:45	
Air Blank		0.000	21:46	
Control Test		0.075	21:47	
Air Blank			21:47	RFI Detect
Air Blank		0.000	21:47	

Data Download Date/Time: 06/30/2017 12:03

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIERNACKI,CHAD,A
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: BIERNACKI,CHAD
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 05/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: DUI
ARREST TIME: 23:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:38

SUBJECT'S LAST NAME: LOWDELL

DATE/TIME OF TEST: 06/02/2017 00:07:18

Diagnostics Check		OK	23:58	
Air Blank		0.000	23:59	
Control Test		0.078	23:59	
Air Blank		0.000	23:59	
Subject Sample	# 1		00:03	No Sample Provided
Breath Volume		0.000	00:03	
Air Blank		0.000	00:03	
Air Blank		0.000	00:05	
Subject Sample	# 2		00:05	Subject Test Refused
Breath Volume		0.000	00:05	
Air Blank		0.000	00:06	
Control Test		0.078	00:06	
Air Blank		0.000	00:07	
Diagnostics Check		OK	00:07	

Data Download Date/Time: 06/30/2017 12:03

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MALOY,JAMES,L
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: N/A,N/A
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 05/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: PROB/PAROL
ARREST TIME: 11:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:40

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 06/03/2017 13:31:34

Diagnostics Check		OK	13:25
Air Blank		0.000	13:26
Control Test		0.078	13:26
Air Blank		0.000	13:26
Subject Sample	# 1	0.000	13:27
Breath Volume		2.132	13:27
Air Blank		0.000	13:27
Air Blank		0.000	13:29
Subject Sample	# 2	0.000	13:30
Breath Volume		2.390	13:30
Air Blank		0.000	13:30
Control Test		0.079	13:30
Air Blank		0.000	13:31
Diagnostics Check		OK	13:31

Data Download Date/Time: 06/30/2017 12:03

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GAYDOS,JONATHAN,A
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: BUSH,SAVANNAH
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 05/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: DUI
ARREST TIME: 17:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:06

SUBJECT'S LAST NAME: HUSS

DATE/TIME OF TEST: 06/05/2017 18:34:22

Diagnostics Check		OK	18:27
Air Blank		0.000	18:28
Control Test		0.078	18:28
Air Blank		0.000	18:29
Subject Sample	# 1	0.165	18:29
Breath Volume		2.113	18:29
Air Blank		0.000	18:30
Air Blank		0.000	18:32
Subject Sample	# 2	0.163	18:32
Breath Volume		1.847	18:32
Air Blank		0.000	18:33
Control Test		0.078	18:33
Air Blank		0.000	18:34
Diagnostics Check		OK	18:34

Data Download Date/Time: 06/30/2017 12:03

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIERNACKI,CHAD,A
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: MOUNT,DEBORAH
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 05/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: DUI
ARREST TIME: 21:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:42

SUBJECT'S LAST NAME: TIFFANY

DATE/TIME OF TEST: 06/09/2017 22:10:37

Diagnostics Check		OK	22:03
Air Blank		0.000	22:04
Control Test		0.078	22:04
Air Blank		0.000	22:05
Subject Sample	# 1	0.149	22:05
Breath Volume		1.257	22:05
Air Blank		0.000	22:06
Air Blank		0.000	22:08
Subject Sample	# 2	0.154	22:09
Breath Volume		1.773	22:09
Air Blank		0.000	22:09
Control Test		0.078	22:10
Air Blank		0.000	22:10
Diagnostics Check		OK	22:10

Data Download Date/Time: 06/30/2017 12:03

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIERNACKI,CHAD,A
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: BIERNACKI,CHAD
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 05/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: DUI
ARREST TIME: 23:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:41

SUBJECT'S LAST NAME: HAWK

DATE/TIME OF TEST: 06/10/2017 00:10:48

Diagnostics Check		OK	00:02
Air Blank		0.000	00:02
Control Test		0.078	00:03
Air Blank		0.000	00:03
Subject Sample	# 1	0.183	00:05
Breath Volume		1.234	00:05
Air Blank		0.000	00:06
Air Blank		0.000	00:07
Subject Sample	# 2	0.186	00:09
Breath Volume		1.246	00:09
Air Blank		0.000	00:09
Control Test		0.075	00:10
Air Blank		0.000	00:10
Diagnostics Check		OK	00:10

Data Download Date/Time: 06/30/2017 12:03

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COOK,DENNIS,K
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: DAY,JASON
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 05/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: ADMIN
ARREST TIME: 12:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:20

SUBJECT'S LAST NAME: COOK

DATE/TIME OF TEST: 06/10/2017 13:01:57

Diagnostics Check		OK	12:55
Air Blank		0.000	12:56
Control Test		0.079	12:56
Air Blank		0.000	12:57
Subject Sample	# 1	0.000	12:57
Breath Volume		4.085	12:57
Air Blank		0.000	12:58
Air Blank		0.000	13:00
Subject Sample	# 2	0.000	13:00
Breath Volume		4.234	13:00
Air Blank		0.000	13:01
Control Test		0.078	13:01
Air Blank		0.000	13:01
Diagnostics Check		OK	13:01

Data Download Date/Time: 06/30/2017 12:03

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MALOY,JAMES,L
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: N/A,N/A
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 05/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: PROB/PAROL
ARREST TIME: 10:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:30

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 06/16/2017 11:25:05

Diagnostics Check		OK	11:19
Air Blank		0.000	11:19
Control Test		0.079	11:20
Air Blank		0.000	11:20
Subject Sample	# 1	0.000	11:20
Breath Volume		2.117	11:20
Air Blank		0.000	11:21
Air Blank		0.000	11:23
Subject Sample	# 2	0.000	11:23
Breath Volume		2.351	11:23
Air Blank		0.000	11:24
Control Test		0.078	11:24
Air Blank		0.000	11:25
Diagnostics Check		OK	11:25

Data Download Date/Time: 06/30/2017 12:03

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHULTZ,JEREMY,J
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: SCHULTZ,JEREMY
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 05/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: DUI
ARREST TIME: 02:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: MCDONALD

DATE/TIME OF TEST: 06/17/2017 03:45:50

Diagnostics Check		OK	03:34	
Air Blank		0.000	03:34	
Control Test		0.078	03:35	
Air Blank		0.000	03:35	
Subject Sample	# 1		03:38	No Sample Provided
Breath Volume		0.000	03:38	
Air Blank		0.000	03:39	
Air Blank		0.000	03:41	
Subject Sample	# 2		03:44	No Sample Provided
Breath Volume		0.000	03:44	
Air Blank		0.000	03:44	
Control Test		0.080	03:45	
Air Blank		0.000	03:45	
Diagnostics Check		OK	03:45	

Data Download Date/Time: 06/30/2017 12:03

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAWLES,COREY,M
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: RAWLES,COREY
ARRESTING OFFICER AGENCY: OTHER County
LAST AGENCY INSPECTION: 05/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: COURT ORD
ARREST TIME: 10:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:24

SUBJECT'S LAST NAME: LANASA

DATE/TIME OF TEST: 06/18/2017 10:50:34

Diagnostics Check		OK	10:44
Air Blank		0.000	10:45
Control Test		0.078	10:45
Air Blank		0.000	10:45
Subject Sample	# 1	0.000	10:46
Breath Volume		2.121	10:46
Air Blank		0.000	10:46
Air Blank		0.000	10:48
Subject Sample	# 2	0.000	10:49
Breath Volume		2.257	10:49
Air Blank		0.000	10:49
Control Test		0.078	10:49
Air Blank		0.000	10:50
Diagnostics Check		OK	10:50

Data Download Date/Time: 06/30/2017 12:04

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIERNACKI,CHAD,A
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: JOINER,JESSE
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 05/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:38

SUBJECT'S LAST NAME: TANNER

DATE/TIME OF TEST: 06/21/2017 03:22:50

Diagnostics Check		OK	03:07	
Air Blank		0.000	03:08	
Control Test		0.078	03:08	
Air Blank		0.000	03:08	
Subject Sample	# 1		03:12	No Sample Provided
Breath Volume		0.000	03:12	
Air Blank		0.000	03:12	
Air Blank		0.000	03:14	
Subject Sample	# 2	0.221	03:15	
Breath Volume		1.812	03:15	
Air Blank		0.000	03:16	
Air Blank		0.000	03:18	
Subject Sample	# 3	0.197	03:21	Volume Not Met
Breath Volume		0.519	03:21	
Air Blank		0.000	03:21	
Control Test		0.077	03:22	
Air Blank		0.000	03:22	
Diagnostics Check		OK	03:22	

Data Download Date/Time: 06/30/2017 12:04

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIERNACKI,CHAD,A
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: JOINER,JESSE
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 05/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:38

SUBJECT'S LAST NAME: TANNER

DATE/TIME OF TEST: 06/21/2017 03:37:44

Diagnostics Check		OK	03:25	
Air Blank		0.000	03:26	
Control Test		0.078	03:26	
Air Blank		0.000	03:27	
Subject Sample	# 1	0.199	03:27	
Breath Volume		2.042	03:27	
Air Blank		0.000	03:28	
Air Blank		0.000	03:30	
Subject Sample	# 2	0.213	03:33	Volume Not Met
Breath Volume		0.781	03:33	
Air Blank		0.000	03:34	
Air Blank		0.000	03:35	
Subject Sample	# 3	0.183	03:36	
Breath Volume		2.097	03:36	
Air Blank		0.000	03:36	
Control Test		0.073	03:37	Control Outside Tolerance
Air Blank		0.000	03:37	

Data Download Date/Time: 06/30/2017 12:04

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCORMICK,CHANTEL,B
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: MCCORMICK,CHANTEL
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 05/31/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: COURT ORD
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:40

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 06/21/2017 16:16:59

Diagnostics Check		OK	16:11
Air Blank		0.000	16:11
Control Test		0.079	16:11
Air Blank		0.000	16:12
Subject Sample	# 1	0.000	16:12
Breath Volume		2.417	16:12
Air Blank		0.000	16:13
Air Blank		0.000	16:15
Subject Sample	# 2	0.000	16:15
Breath Volume		2.644	16:15
Air Blank		0.000	16:16
Control Test		0.077	16:16
Air Blank		0.000	16:16
Diagnostics Check		OK	16:16

Data Download Date/Time: 06/30/2017 12:04

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,CURTIS,C
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: BUSH,SAVANNAH
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 05/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: COURT ORD
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:35

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 06/28/2017 18:05:15

Diagnostics Check		OK	17:59
Air Blank		0.000	17:59
Control Test		0.078	18:00
Air Blank		0.000	18:00
Subject Sample	# 1	0.000	18:00
Breath Volume		2.414	18:00
Air Blank		0.000	18:01
Air Blank		0.000	18:03
Subject Sample	# 2	0.000	18:03
Breath Volume		3.886	18:03
Air Blank		0.000	18:04
Control Test		0.077	18:04
Air Blank		0.000	18:05
Diagnostics Check		OK	18:05

Data Download Date/Time: 06/30/2017 12:04

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001320

INSTRUMENT REGISTERED TO: OKEECHOBEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POLLOCK,STEVEN,D
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: POLLOCK,STEVEN
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: DUI
ARREST TIME: 22:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:14

SUBJECT'S LAST NAME: MORELAND

DATE/TIME OF TEST: 06/02/2017 23:40:49

Diagnostics Check		OK	23:34
Air Blank		0.000	23:35
Control Test		0.080	23:35
Air Blank		0.000	23:35
Subject Sample	# 1	0.000	23:36
Breath Volume		4.500	23:36
Air Blank		0.000	23:36
Air Blank		0.000	23:38
Subject Sample	# 2	0.000	23:39
Breath Volume		4.636	23:39
Air Blank		0.000	23:39
Control Test		0.080	23:40
Air Blank		0.000	23:40
Diagnostics Check		OK	23:40

Data Download Date/Time: 07/07/2017 09:57

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001320

INSTRUMENT REGISTERED TO: OKEECHOBEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POLLOCK,STEVEN,D
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: POLLOCK,STEVEN
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: DUI
ARREST TIME: 17:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:42

SUBJECT'S LAST NAME: BASULTO

DATE/TIME OF TEST: 06/04/2017 18:05:28

Diagnostics Check		OK	18:02	
Air Blank		0.000	18:02	
Control Test		0.080	18:03	
Air Blank		0.000	18:03	
Subject Sample	# 1		18:04	Subject Test Refused
Breath Volume		0.000	18:04	
Air Blank		0.000	18:04	
Control Test		0.080	18:04	
Air Blank		0.000	18:05	
Diagnostics Check		OK	18:05	

Data Download Date/Time: 07/07/2017 09:57

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001320

INSTRUMENT REGISTERED TO: OKEECHOBEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POLLOCK,STEVEN,D
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: POLLOCK,STEVEN
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: DUI
ARREST TIME: 22:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:41

SUBJECT'S LAST NAME: LINDSLEY

DATE/TIME OF TEST: 06/16/2017 23:06:48

Diagnostics Check		OK	23:02	
Air Blank		0.000	23:03	
Control Test		0.080	23:03	
Air Blank		0.000	23:03	
Subject Sample	# 1		23:05	Subject Test Refused
Breath Volume		0.000	23:05	
Air Blank		0.000	23:05	
Control Test		0.080	23:06	
Air Blank		0.000	23:06	
Diagnostics Check		OK	23:06	

Data Download Date/Time: 07/07/2017 09:57

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001320

INSTRUMENT REGISTERED TO: OKEECHOBEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POLLOCK,STEVEN,D
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: POLLOCK,STEVEN
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: DUI
ARREST TIME: 01:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:31

SUBJECT'S LAST NAME: NETTLE JR

DATE/TIME OF TEST: 06/19/2017 02:16:37

Diagnostics Check		OK	02:13	
Air Blank		0.000	02:14	
Control Test		0.080	02:14	
Air Blank		0.000	02:14	
Subject Sample	# 1		02:15	Subject Test Refused
Breath Volume		0.000	02:15	
Air Blank		0.000	02:15	
Control Test		0.080	02:16	
Air Blank		0.000	02:16	
Diagnostics Check		OK	02:16	

Data Download Date/Time: 07/07/2017 09:57

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001320

INSTRUMENT REGISTERED TO: OKEECHOBEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROGERS,DAVID,E
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: SATALLANTE,DEVON
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: DUI
ARREST TIME: 18:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:03

SUBJECT'S LAST NAME: PATEL

DATE/TIME OF TEST: 06/28/2017 19:35:21

Diagnostics Check		OK	19:23	
Air Blank		0.000	19:24	
Control Test		0.080	19:24	
Air Blank		0.000	19:25	
Subject Sample	# 1	0.158	19:28	Volume Not Met
Breath Volume		0.246	19:28	
Air Blank		0.000	19:28	
Air Blank		0.000	19:30	
Subject Sample	# 2	0.152	19:33	Volume Not Met
Breath Volume		0.273	19:33	
Air Blank		0.000	19:34	
Control Test		0.079	19:34	
Air Blank		0.000	19:35	
Diagnostics Check		OK	19:35	

Data Download Date/Time: 07/07/2017 09:57

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUZMAN,MELISSA,M
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: RUIZ,MICHAEL
ARRESTING OFFICER AGENCY: Fellsmere PD
LAST AGENCY INSPECTION: 05/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 00:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:44

SUBJECT'S LAST NAME: GOMEZ-SANCHEZ

DATE/TIME OF TEST: 06/02/2017 01:07:57

Diagnostics Check		OK	01:04	
Air Blank		0.000	01:05	
Control Test		0.082	01:05	
Air Blank		0.000	01:06	
Subject Sample	# 1		01:06	RFI Detect
Breath Volume		0.000	01:06	
Air Blank		0.000	01:07	
Control Test		0.081	01:07	
Air Blank		0.000	01:07	
Diagnostics Check		OK	01:07	

Data Download Date/Time: 06/27/2017 05:47

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUZMAN,MELISSA,M
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: RUIZ,MICHAEL
ARRESTING OFFICER AGENCY: Fellsmere PD
LAST AGENCY INSPECTION: 05/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 00:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:44

SUBJECT'S LAST NAME: GOMEZ-SANCHEZ

DATE/TIME OF TEST: 06/02/2017 01:15:30

Diagnostics Check		OK	01:09
Air Blank		0.000	01:09
Control Test		0.081	01:10
Air Blank		0.000	01:10
Subject Sample	# 1	0.212	01:11
Breath Volume		1.582	01:11
Air Blank		0.000	01:11
Air Blank		0.000	01:13
Subject Sample	# 2	0.201	01:13
Breath Volume		1.796	01:13
Air Blank		0.000	01:14
Control Test		0.081	01:14
Air Blank		0.000	01:15
Diagnostics Check		OK	01:15

Data Download Date/Time: 06/27/2017 05:47

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOLSKI,JEFFREY,R
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: JAWORSKI,KEVIN
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 05/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 17:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:20

SUBJECT'S LAST NAME: BORKOWSKI

DATE/TIME OF TEST: 06/02/2017 17:55:42

Diagnostics Check		OK	17:44	
Air Blank		0.000	17:44	
Control Test		0.081	17:45	
Air Blank		0.000	17:45	
Subject Sample	# 1		17:48	No Sample Provided
Breath Volume		0.000	17:48	
Air Blank		0.000	17:49	
Air Blank		0.000	17:51	
Subject Sample	# 2		17:54	No Sample Provided
Breath Volume		0.000	17:54	
Air Blank		0.000	17:54	
Control Test		0.080	17:55	
Air Blank		0.000	17:55	
Diagnostics Check		OK	17:55	

Data Download Date/Time: 06/27/2017 05:47

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOLSKI,JEFFREY,R
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: JAWORSKI,K
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 05/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 17:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:20

SUBJECT'S LAST NAME: BORKOWSKI

DATE/TIME OF TEST: 06/02/2017 18:10:40

Diagnostics Check		OK	18:07	
Air Blank		0.000	18:08	
Control Test		0.081	18:08	
Air Blank		0.000	18:09	
Subject Sample	# 1		18:09	Subject Test Refused
Breath Volume		0.000	18:09	
Air Blank		0.000	18:09	
Control Test		0.080	18:10	
Air Blank		0.000	18:10	
Diagnostics Check		OK	18:10	

Data Download Date/Time: 06/27/2017 05:47

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,QUINTON,T
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: JAWORSKI,KEVIN
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 05/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 22:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:15

SUBJECT'S LAST NAME: COOPER

DATE/TIME OF TEST: 06/04/2017 22:45:15

Diagnostics Check		OK	22:36	
Air Blank		0.000	22:37	
Control Test		0.082	22:37	
Air Blank		0.000	22:38	
Subject Sample	# 1		22:41	No Sample Provided
Breath Volume		0.000	22:41	
Air Blank		0.000	22:41	
Air Blank		0.000	22:43	
Subject Sample	# 2		22:43	Subject Test Refused
Breath Volume		0.000	22:43	
Air Blank		0.000	22:44	
Control Test		0.082	22:44	
Air Blank		0.000	22:45	
Diagnostics Check		OK	22:45	

Data Download Date/Time: 06/27/2017 05:47

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SWINDELL,CHRISTOPHER,E
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: FARLESS,GARY
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 05/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 23:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:12

SUBJECT'S LAST NAME: ROSE

DATE/TIME OF TEST: 06/09/2017 00:36:06

Diagnostics Check		OK	00:33	
Air Blank		0.000	00:33	
Control Test		0.081	00:34	
Air Blank		0.000	00:34	
Subject Sample	# 1		00:34	Subject Test Refused
Breath Volume		0.000	00:34	
Air Blank		0.000	00:35	
Control Test		0.081	00:35	
Air Blank		0.000	00:36	
Diagnostics Check		OK	00:36	

Data Download Date/Time: 06/27/2017 05:47

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SWINDELL,CHRISTOPHER,E
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: RUIZ,MIKE
ARRESTING OFFICER AGENCY: Fellsmere PD
LAST AGENCY INSPECTION: 05/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 01:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:22

SUBJECT'S LAST NAME: DE COSA

DATE/TIME OF TEST: 06/10/2017 02:50:36

Diagnostics Check		OK	02:44
Air Blank		0.000	02:44
Control Test		0.081	02:44
Air Blank		0.000	02:45
Subject Sample	# 1	0.103	02:46
Breath Volume		2.253	02:46
Air Blank		0.000	02:46
Air Blank		0.000	02:48
Subject Sample	# 2	0.103	02:49
Breath Volume		1.792	02:49
Air Blank		0.000	02:49
Control Test		0.081	02:50
Air Blank		0.000	02:50
Diagnostics Check		OK	02:50

Data Download Date/Time: 06/27/2017 05:47

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROLLE,MELINDA,M
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: PEREZ,CRISTAL
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 05/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 17:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:40

SUBJECT'S LAST NAME: SEVERINO

DATE/TIME OF TEST: 06/10/2017 18:38:01

Diagnostics Check		OK	18:35	
Air Blank		0.000	18:35	
Control Test		0.081	18:35	
Air Blank		0.000	18:36	
Subject Sample	# 1		18:36	Subject Test Refused
Breath Volume		0.000	18:36	
Air Blank		0.000	18:37	
Control Test		0.081	18:37	
Air Blank		0.000	18:37	
Diagnostics Check		OK	18:38	

Data Download Date/Time: 06/27/2017 05:47

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALLAGHER,MATTHEW,T
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: GALLAGHER,MATTHEW
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 05/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 01:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: PERDUE

DATE/TIME OF TEST: 06/11/2017 02:19:48

Diagnostics Check		OK	02:16	
Air Blank		0.000	02:16	
Control Test		0.081	02:16	
Air Blank		0.000	02:17	
Subject Sample	# 1		02:18	Subject Test Refused
Breath Volume		0.000	02:18	
Air Blank		0.000	02:18	
Control Test		0.081	02:19	
Air Blank		0.000	02:19	
Diagnostics Check		OK	02:19	

Data Download Date/Time: 06/27/2017 05:47

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,LEE,A
BREATH TEST OPERATOR AGENCY: Vero Beach PD
ARRESTING OFFICER: EVANS,LEE
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 05/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 02:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:26

SUBJECT'S LAST NAME: BURRIS

DATE/TIME OF TEST: 06/11/2017 03:05:02

Diagnostics Check		OK	03:02	
Air Blank		0.000	03:02	
Control Test		0.081	03:03	
Air Blank		0.000	03:03	
Subject Sample	# 1		03:03	Subject Test Refused
Breath Volume		0.000	03:03	
Air Blank		0.000	03:04	
Control Test		0.080	03:04	
Air Blank		0.000	03:04	
Diagnostics Check		OK	03:05	

Data Download Date/Time: 06/27/2017 05:47

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,TERRENCE,R
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: KLASSEN,BRYAN
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 05/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 16:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:11

SUBJECT'S LAST NAME: MAINE

DATE/TIME OF TEST: 06/11/2017 18:40:34

Diagnostics Check		OK	18:37	
Air Blank		0.000	18:38	
Control Test		0.081	18:38	
Air Blank		0.000	18:38	
Subject Sample	# 1		18:39	Subject Test Refused
Breath Volume		0.000	18:39	
Air Blank		0.000	18:39	
Control Test		0.081	18:39	
Air Blank		0.000	18:40	
Diagnostics Check		OK	18:40	

Data Download Date/Time: 06/27/2017 05:47

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMILTON,DEVIN,C
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: BUFFINGTON,BEN
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 05/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 02:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:13

SUBJECT'S LAST NAME: MERCER

DATE/TIME OF TEST: 06/13/2017 03:42:02

Diagnostics Check		OK	03:33
Air Blank		0.000	03:33
Control Test		0.081	03:34
Air Blank		0.000	03:34
Subject Sample	# 1	0.184	03:37
Breath Volume		2.199	03:37
Air Blank		0.000	03:37
Air Blank		0.000	03:39
Subject Sample	# 2	0.183	03:40
Breath Volume		2.113	03:40
Air Blank		0.000	03:41
Control Test		0.081	03:41
Air Blank		0.000	03:41
Diagnostics Check		OK	03:42

Data Download Date/Time: 06/27/2017 05:47

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SWINDELL,CHRISTOPHER,E
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: FARINACCI,DYLAN
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 05/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 00:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:44

SUBJECT'S LAST NAME: DAWKINS

DATE/TIME OF TEST: 06/18/2017 01:18:18

Diagnostics Check		OK	01:15	
Air Blank		0.000	01:15	
Control Test		0.082	01:16	
Air Blank		0.000	01:16	
Subject Sample	# 1		01:16	Subject Test Refused
Breath Volume		0.000	01:16	
Air Blank		0.000	01:17	
Control Test		0.082	01:17	
Air Blank		0.000	01:18	
Diagnostics Check		OK	01:18	

Data Download Date/Time: 06/27/2017 05:47

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOLSKI,JEFFREY,R
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MATTHEWS,G
ARRESTING OFFICER AGENCY: Sebastian PD
LAST AGENCY INSPECTION: 05/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 22:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:33

SUBJECT'S LAST NAME: WILCKEN

DATE/TIME OF TEST: 06/18/2017 23:58:24

Diagnostics Check		OK	23:54	
Air Blank		0.000	23:54	
Control Test		0.080	23:55	
Air Blank		0.000	23:55	
Subject Sample	# 1		23:57	Subject Test Refused
Breath Volume		0.000	23:57	
Air Blank		0.000	23:57	
Control Test		0.080	23:57	
Air Blank		0.000	23:58	
Diagnostics Check		OK	23:58	

Data Download Date/Time: 06/27/2017 05:47

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUZMAN,MELISSA,M
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: DALEY,SHAWN
ARRESTING OFFICER AGENCY: Sebastian PD
LAST AGENCY INSPECTION: 05/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 00:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:44

SUBJECT'S LAST NAME: TUCKER

DATE/TIME OF TEST: 06/20/2017 01:13:34

Diagnostics Check		OK	01:07
Air Blank		0.000	01:07
Control Test		0.082	01:08
Air Blank		0.000	01:08
Subject Sample	# 1	0.217	01:09
Breath Volume		1.628	01:09
Air Blank		0.000	01:09
Air Blank		0.000	01:11
Subject Sample	# 2	0.211	01:12
Breath Volume		2.031	01:12
Air Blank		0.000	01:12
Control Test		0.080	01:12
Air Blank		0.000	01:13
Diagnostics Check		OK	01:13

Data Download Date/Time: 06/27/2017 05:47

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRIMAVERA,SEAN,M
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: RISPOLI,MICHELLE
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 05/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 00:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:51

SUBJECT'S LAST NAME: GAMEZ CORTEZ

DATE/TIME OF TEST: 06/24/2017 01:34:08

Diagnostics Check		OK	01:31	
Air Blank		0.000	01:31	
Control Test		0.081	01:32	
Air Blank		0.000	01:32	
Subject Sample	# 1		01:32	Subject Test Refused
Breath Volume		0.000	01:32	
Air Blank		0.000	01:33	
Control Test		0.082	01:33	
Air Blank		0.000	01:34	
Diagnostics Check		OK	01:34	

Data Download Date/Time: 06/27/2017 05:47

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KANE,JAMES,G
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: JAWORSKI,KEVIN
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 05/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 12:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:58

SUBJECT'S LAST NAME: MAUGERI

DATE/TIME OF TEST: 06/24/2017 13:27:54

Diagnostics Check		OK	13:21
Air Blank		0.000	13:22
Control Test		0.082	13:22
Air Blank		0.000	13:23
Subject Sample	# 1	0.000	13:23
Breath Volume		1.988	13:23
Air Blank		0.000	13:24
Air Blank		0.000	13:26
Subject Sample	# 2	0.000	13:26
Breath Volume		2.269	13:26
Air Blank		0.000	13:27
Control Test		0.081	13:27
Air Blank		0.000	13:27
Diagnostics Check		OK	13:27

Data Download Date/Time: 06/27/2017 05:47

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MAEGAN,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: BUFFINGTON,BEN
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 05/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 17:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:45

SUBJECT'S LAST NAME: FRANCIS

DATE/TIME OF TEST: 06/26/2017 18:28:38

Diagnostics Check		OK	18:22
Air Blank		0.000	18:22
Control Test		0.082	18:22
Air Blank		0.000	18:23
Subject Sample	# 1	0.163	18:23
Breath Volume		1.808	18:23
Air Blank		0.000	18:24
Air Blank		0.000	18:26
Subject Sample	# 2	0.175	18:27
Breath Volume		2.058	18:27
Air Blank		0.000	18:27
Control Test		0.081	18:28
Air Blank		0.000	18:28
Diagnostics Check		OK	18:28

Data Download Date/Time: 06/27/2017 05:47

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KANTORSKI,MARK,A
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: SEBRIS,ANDREW
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 06/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 00:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: ZIMMER

DATE/TIME OF TEST: 06/30/2017 01:14:38

Diagnostics Check		OK	01:08
Air Blank		0.000	01:08
Control Test		0.082	01:08
Air Blank		0.000	01:09
Subject Sample	# 1	0.118	01:09
Breath Volume		2.335	01:09
Air Blank		0.000	01:10
Air Blank		0.000	01:12
Subject Sample	# 2	0.115	01:13
Breath Volume		2.472	01:13
Air Blank		0.000	01:13
Control Test		0.081	01:14
Air Blank		0.000	01:14
Diagnostics Check		OK	01:14

Data Download Date/Time: 07/21/2017 06:04

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001341

INSTRUMENT REGISTERED TO: DESOTO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 05/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:04

SUBJECT'S LAST NAME: FLEMING

DATE/TIME OF TEST: 06/10/2017 01:31:40

Diagnostics Check		OK	01:25
Air Blank		0.000	01:25
Control Test		0.080	01:26
Air Blank		0.000	01:26
Subject Sample	# 1	0.118	01:27
Breath Volume		2.019	01:27
Air Blank		0.000	01:27
Air Blank		0.000	01:29
Subject Sample	# 2	0.103	01:30
Breath Volume		1.332	01:30
Air Blank		0.000	01:30
Control Test		0.079	01:31
Air Blank		0.000	01:31
Diagnostics Check		OK	01:31

Data Download Date/Time: 06/27/2017 10:19

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001341

INSTRUMENT REGISTERED TO: DESOTO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HORTON,EDMOND,W
BREATH TEST OPERATOR AGENCY: DeSoto County SO
ARRESTING OFFICER: QUIGLEY,MEGAN
ARRESTING OFFICER AGENCY: DeSoto County SO
LAST AGENCY INSPECTION: 05/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:42

SUBJECT'S LAST NAME: HAMPTON

DATE/TIME OF TEST: 06/10/2017 23:12:24

Diagnostics Check		OK	23:06
Air Blank		0.000	23:06
Control Test		0.080	23:07
Air Blank		0.000	23:07
Subject Sample	# 1	0.125	23:07
Breath Volume		1.519	23:07
Air Blank		0.000	23:08
Air Blank		0.000	23:10
Subject Sample	# 2	0.128	23:10
Breath Volume		2.339	23:10
Air Blank		0.000	23:11
Control Test		0.079	23:11
Air Blank		0.000	23:12
Diagnostics Check		OK	23:12

Data Download Date/Time: 06/27/2017 10:19

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001341

INSTRUMENT REGISTERED TO: DESOTO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KERR,CALEB,W
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: KERR,CALEB
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 05/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: LOCKLEAR

DATE/TIME OF TEST: 06/11/2017 01:55:08

Diagnostics Check		OK	01:52	
Air Blank		0.000	01:52	
Control Test		0.079	01:53	
Air Blank		0.000	01:53	
Subject Sample	# 1		01:53	Subject Test Refused
Breath Volume		0.000	01:53	
Air Blank		0.000	01:54	
Control Test		0.079	01:54	
Air Blank		0.000	01:55	
Diagnostics Check		OK	01:55	

Data Download Date/Time: 06/27/2017 10:19

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001341

INSTRUMENT REGISTERED TO: DESOTO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALEXANDER,DAVID,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: ALEXANDER,DAVID
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 05/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: RODRIQUEZ

DATE/TIME OF TEST: 06/11/2017 02:43:02

Diagnostics Check		OK	02:40	
Air Blank		0.000	02:40	
Control Test		0.079	02:40	
Air Blank		0.000	02:41	
Subject Sample	# 1		02:41	Subject Test Refused
Breath Volume		0.000	02:41	
Air Blank		0.000	02:42	
Control Test		0.079	02:42	
Air Blank		0.000	02:42	
Diagnostics Check		OK	02:43	

Data Download Date/Time: 06/27/2017 10:19

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001341

INSTRUMENT REGISTERED TO: DESOTO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAUDILL,CHERI,L
BREATH TEST OPERATOR AGENCY: DeSoto County SO
ARRESTING OFFICER: LAMBKA,JORDON
ARRESTING OFFICER AGENCY: Arcadia PD
LAST AGENCY INSPECTION: 05/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: MULLET

DATE/TIME OF TEST: 06/22/2017 02:29:35

Diagnostics Check		OK	02:26	
Air Blank		0.000	02:27	
Control Test		0.079	02:27	
Air Blank		0.000	02:28	
Subject Sample	# 1		02:28	Subject Test Refused
Breath Volume		0.000	02:28	
Air Blank		0.000	02:28	
Control Test		0.080	02:29	
Air Blank		0.000	02:29	
Diagnostics Check		OK	02:29	

Data Download Date/Time: 06/27/2017 10:19

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001341

INSTRUMENT REGISTERED TO: DESOTO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DITTMAN,LAVERNE,T
BREATH TEST OPERATOR AGENCY: DeSoto County SO
ARRESTING OFFICER: LAMBERT,CASEY
ARRESTING OFFICER AGENCY: Arcadia PD
LAST AGENCY INSPECTION: 05/25/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 18:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:05

SUBJECT'S LAST NAME: PAWELL

DATE/TIME OF TEST: 06/24/2017 19:39:12

Diagnostics Check		OK	19:30
Air Blank		0.000	19:30
Control Test		0.080	19:31
Air Blank		0.000	19:31
Subject Sample	# 1	0.322	19:31
Breath Volume		2.347	19:31
Air Blank		0.000	19:32
Air Blank		0.000	19:34
Subject Sample	# 2	0.344	19:34
Breath Volume		2.710	19:34
Air Blank		0.000	19:35
Air Blank		0.000	19:37
Subject Sample	# 3	0.320	19:37
Breath Volume		3.179	19:37
Air Blank		0.000	19:38
Control Test		0.079	19:38
Air Blank		0.000	19:39
Diagnostics Check		OK	19:39

Data Download Date/Time: 06/27/2017 10:19

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001341

INSTRUMENT REGISTERED TO: DESOTO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEARSON,DEANN,M
BREATH TEST OPERATOR AGENCY: DeSoto County SO
ARRESTING OFFICER: MILEY,CARL
ARRESTING OFFICER AGENCY: DeSoto County SO
LAST AGENCY INSPECTION: 06/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 15:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:33

SUBJECT'S LAST NAME: STEYER

DATE/TIME OF TEST: 06/28/2017 16:00:40

Diagnostics Check		OK	15:53	
Air Blank		0.000	15:54	
Control Test		0.080	15:54	
Air Blank		0.000	15:55	
Subject Sample	# 1	0.232	15:56	
Breath Volume		1.363	15:56	
Air Blank		0.000	15:56	
Air Blank		0.000	15:58	
Subject Sample	# 2	0.261	15:59	
Breath Volume		1.160	15:59	
Air Blank			15:59	Improper Sample
Air Blank		0.000	15:59	
Control Test		0.080	16:00	
Air Blank		0.000	16:00	
Diagnostics Check		OK	16:00	

Data Download Date/Time: 07/21/2017 08:22

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001341

INSTRUMENT REGISTERED TO: DESOTO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEARSON,DEANN,M
BREATH TEST OPERATOR AGENCY: DeSoto County SO
ARRESTING OFFICER: MILEY,CARL
ARRESTING OFFICER AGENCY: DeSoto County SO
LAST AGENCY INSPECTION: 06/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 15:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:33

SUBJECT'S LAST NAME: STEYER

DATE/TIME OF TEST: 06/28/2017 16:25:59

Diagnostics Check		OK	16:09	
Air Blank		0.000	16:09	
Control Test		0.080	16:10	
Air Blank		0.000	16:10	
Subject Sample	# 1	0.253	16:13	
Breath Volume		1.230	16:13	
Air Blank		0.000	16:13	
Air Blank		0.000	16:15	
Subject Sample	# 2	0.213	16:18	Volume Not Met
Breath Volume		0.792	16:18	
Air Blank		0.000	16:19	
Air Blank		0.000	16:21	
Subject Sample	# 3	0.238	16:24	Volume Not Met
Breath Volume		0.203	16:24	
Air Blank		0.000	16:25	
Control Test		0.077	16:25	
Air Blank		0.000	16:25	
Diagnostics Check		OK	16:25	

Data Download Date/Time: 07/21/2017 08:22

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001341

INSTRUMENT REGISTERED TO: DESOTO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FILIPOWICZ,KEVIN,M
BREATH TEST OPERATOR AGENCY: DeSoto County SO
ARRESTING OFFICER: PACHECO,ERIQ
ARRESTING OFFICER AGENCY: DeSoto County SO
LAST AGENCY INSPECTION: 06/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 11:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:22

SUBJECT'S LAST NAME: SKINNER

DATE/TIME OF TEST: 06/30/2017 11:58:38

Diagnostics Check		OK	11:47	
Air Blank		0.000	11:47	
Control Test		0.079	11:47	
Air Blank		0.000	11:48	
Subject Sample	# 1	0.213	11:51	Volume Not Met
Breath Volume		0.226	11:51	
Air Blank		0.000	11:52	
Air Blank		0.000	11:53	
Subject Sample	# 2	0.178	11:57	Volume Not Met
Breath Volume		0.035	11:57	
Air Blank		0.000	11:57	
Control Test		0.079	11:58	
Air Blank		0.000	11:58	
Diagnostics Check		OK	11:58	

Data Download Date/Time: 07/21/2017 08:22

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001341

INSTRUMENT REGISTERED TO: DESOTO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FILIPOWICZ,KEVIN,M
BREATH TEST OPERATOR AGENCY: DeSoto County SO
ARRESTING OFFICER: PACHECO,ERIQ
ARRESTING OFFICER AGENCY: DeSoto County SO
LAST AGENCY INSPECTION: 06/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 11:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:22

SUBJECT'S LAST NAME: SKINNER

DATE/TIME OF TEST: 06/30/2017 12:12:30

Diagnostics Check		OK	12:01	
Air Blank		0.000	12:01	
Control Test		0.078	12:01	
Air Blank		0.000	12:02	
Subject Sample	# 1	0.191	12:05	Volume Not Met
Breath Volume		0.277	12:05	
Air Blank		0.000	12:06	
Air Blank		0.000	12:08	
Subject Sample	# 2		12:10	Subject Test Refused
Breath Volume		0.019	12:10	
Air Blank		0.000	12:11	
Control Test		0.078	12:11	
Air Blank		0.000	12:12	
Diagnostics Check		OK	12:12	

Data Download Date/Time: 07/21/2017 08:22

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PETERSON,STEFANIE,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: PETERSON,STEFANIE
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 05/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:55

SUBJECT'S LAST NAME: FERNANDEZ MONTANO

DATE/TIME OF TEST: 06/02/2017 21:28:30

Diagnostics Check		OK	21:16	
Air Blank		0.000	21:16	
Control Test		0.077	21:16	
Air Blank		0.000	21:17	
Subject Sample	# 1	0.161	21:17	
Breath Volume		1.546	21:17	
Air Blank		0.000	21:18	
Air Blank		0.000	21:20	
Subject Sample	# 2	0.142	21:23	Volume Not Met
Breath Volume		0.281	21:23	
Air Blank		0.000	21:24	
Air Blank		0.000	21:26	
Subject Sample	# 3	0.156	21:26	
Breath Volume		1.445	21:26	
Air Blank		0.000	21:27	
Control Test		0.078	21:27	
Air Blank		0.000	21:28	
Diagnostics Check		OK	21:28	

Data Download Date/Time: 06/28/2017 07:30

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NELSON,HALSTON,L
BREATH TEST OPERATOR AGENCY: North Port PD
ARRESTING OFFICER: NELSON,HALSTON
ARRESTING OFFICER AGENCY: North Port PD
LAST AGENCY INSPECTION: 05/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:02

SUBJECT'S LAST NAME: SPITTLER

DATE/TIME OF TEST: 06/02/2017 21:46:16

Diagnostics Check		OK	21:39
Air Blank		0.000	21:40
Control Test		0.077	21:40
Air Blank		0.000	21:41
Subject Sample	# 1	0.268	21:41
Breath Volume		2.738	21:41
Air Blank		0.000	21:42
Air Blank		0.000	21:44
Subject Sample	# 2	0.256	21:44
Breath Volume		4.183	21:44
Air Blank		0.000	21:45
Control Test		0.076	21:45
Air Blank		0.000	21:46
Diagnostics Check		OK	21:46

Data Download Date/Time: 06/28/2017 07:30

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NAMORS,JASON,R
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: TURNER,B
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 05/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:03

SUBJECT'S LAST NAME: WARE

DATE/TIME OF TEST: 06/04/2017 02:40:52

Diagnostics Check		OK	02:33
Air Blank		0.000	02:33
Control Test		0.079	02:34
Air Blank		0.000	02:34
Subject Sample	# 1	0.170	02:35
Breath Volume		3.238	02:35
Air Blank		0.000	02:36
Air Blank		0.000	02:38
Subject Sample	# 2	0.162	02:39
Breath Volume		3.226	02:39
Air Blank		0.000	02:39
Control Test		0.079	02:40
Air Blank		0.000	02:40
Diagnostics Check		OK	02:40

Data Download Date/Time: 06/28/2017 07:30

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KERR,CALEB,W
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: KERR,CALEB
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 05/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: WEAVER

DATE/TIME OF TEST: 06/04/2017 02:54:20

Diagnostics Check		OK	02:47
Air Blank		0.000	02:48
Control Test		0.079	02:48
Air Blank		0.000	02:48
Subject Sample	# 1	0.161	02:49
Breath Volume		1.406	02:49
Air Blank		0.000	02:50
Air Blank		0.000	02:52
Subject Sample	# 2	0.175	02:52
Breath Volume		1.617	02:52
Air Blank		0.000	02:53
Control Test		0.079	02:53
Air Blank		0.000	02:54
Diagnostics Check		OK	02:54

Data Download Date/Time: 06/28/2017 07:30

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PUFAHL,G,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: KOSTYAL,M
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 05/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:42

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 06/04/2017 03:09:47

Diagnostics Check		OK	03:02
Air Blank		0.000	03:02
Control Test		0.079	03:03
Air Blank		0.000	03:03
Subject Sample	# 1	0.000	03:05
Breath Volume		3.351	03:05
Air Blank		0.000	03:05
Air Blank		0.000	03:07
Subject Sample	# 2	0.000	03:08
Breath Volume		3.523	03:08
Air Blank		0.000	03:08
Control Test		0.080	03:09
Air Blank		0.000	03:09
Diagnostics Check		OK	03:09

Data Download Date/Time: 06/28/2017 07:30

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PUFAHL,G,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: MARTIN,S
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 05/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: CORNELIUS

DATE/TIME OF TEST: 06/04/2017 03:34:39

Diagnostics Check		OK	03:28
Air Blank		0.000	03:28
Control Test		0.079	03:29
Air Blank		0.000	03:29
Subject Sample	# 1	0.148	03:30
Breath Volume		1.585	03:30
Air Blank		0.000	03:30
Air Blank		0.000	03:32
Subject Sample	# 2	0.148	03:33
Breath Volume		1.542	03:33
Air Blank		0.000	03:33
Control Test		0.079	03:34
Air Blank		0.000	03:34
Diagnostics Check		OK	03:34

Data Download Date/Time: 06/28/2017 07:30

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURLEY,SHAUN,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: SCHUMANN,D
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 05/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 18:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:20

SUBJECT'S LAST NAME: KNAPCZYK

DATE/TIME OF TEST: 06/07/2017 20:03:05

Diagnostics Check		OK	19:47	
Air Blank		0.000	19:47	
Control Test		0.080	19:48	
Air Blank		0.000	19:48	
Subject Sample	# 1	0.203	19:51	Volume Not Met
Breath Volume		0.074	19:51	
Air Blank		0.000	19:52	
Air Blank		0.000	19:54	
Subject Sample	# 2	0.212	19:55	
Breath Volume		1.246	19:55	
Air Blank		0.000	19:56	
Air Blank		0.000	19:58	
Subject Sample	# 3	0.176	20:01	Volume Not Met
Breath Volume		0.414	20:01	
Air Blank		0.000	20:02	
Control Test		0.077	20:02	
Air Blank		0.000	20:03	
Diagnostics Check		OK	20:03	

Data Download Date/Time: 06/28/2017 07:30

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NELSON,HALSTON,L
BREATH TEST OPERATOR AGENCY: North Port PD
ARRESTING OFFICER: NELSON,HALSTON
ARRESTING OFFICER AGENCY: North Port PD
LAST AGENCY INSPECTION: 05/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: ROSS

DATE/TIME OF TEST: 06/08/2017 05:10:52

Diagnostics Check		OK	05:03
Air Blank		0.000	05:04
Control Test		0.080	05:04
Air Blank		0.000	05:04
Subject Sample	# 1	0.141	05:06
Breath Volume		2.261	05:06
Air Blank		0.000	05:06
Air Blank		0.000	05:08
Subject Sample	# 2	0.140	05:09
Breath Volume		2.828	05:09
Air Blank		0.000	05:09
Control Test		0.078	05:10
Air Blank		0.000	05:10
Diagnostics Check		OK	05:10

Data Download Date/Time: 06/28/2017 07:30

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANDERSON,DEPUTY,C
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: POINSETT,DEPUTY
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 05/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:11

SUBJECT'S LAST NAME: ANGIER

DATE/TIME OF TEST: 06/09/2017 01:47:36

Diagnostics Check		OK	01:41
Air Blank		0.000	01:41
Control Test		0.080	01:41
Air Blank		0.000	01:42
Subject Sample	# 1	0.324	01:42
Breath Volume		2.375	01:42
Air Blank		0.000	01:43
Air Blank		0.000	01:45
Subject Sample	# 2	0.323	01:45
Breath Volume		2.621	01:45
Air Blank		0.000	01:46
Control Test		0.079	01:47
Air Blank		0.000	01:47
Diagnostics Check		OK	01:47

Data Download Date/Time: 06/28/2017 07:30

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZCOLON,DANIEL,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: SCHUMANN,DEPUTY
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 05/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 18:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:30

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 06/09/2017 20:03:59

Diagnostics Check		OK	19:57
Air Blank		0.000	19:58
Control Test		0.081	19:58
Air Blank		0.000	19:59
Subject Sample	# 1	0.000	19:59
Breath Volume		4.164	19:59
Air Blank		0.000	20:00
Air Blank		0.000	20:02
Subject Sample	# 2	0.000	20:02
Breath Volume		4.519	20:02
Air Blank		0.000	20:03
Control Test		0.081	20:03
Air Blank		0.000	20:03
Diagnostics Check		OK	20:03

Data Download Date/Time: 06/28/2017 07:30

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUBRANO,DEP,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: PERRY,N
ARRESTING OFFICER AGENCY: Venice PD
LAST AGENCY INSPECTION: 05/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:18

SUBJECT'S LAST NAME: HOZA

DATE/TIME OF TEST: 06/10/2017 00:52:30

Diagnostics Check		OK	00:46
Air Blank		0.000	00:46
Control Test		0.080	00:46
Air Blank		0.000	00:47
Subject Sample	# 1	0.000	00:48
Breath Volume		2.726	00:48
Air Blank		0.000	00:48
Air Blank		0.000	00:50
Subject Sample	# 2	0.000	00:51
Breath Volume		3.097	00:51
Air Blank		0.000	00:51
Control Test		0.080	00:51
Air Blank		0.000	00:52
Diagnostics Check		OK	00:52

Data Download Date/Time: 06/28/2017 07:30

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHAETZER,DEPUTY,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: PISCORIK,OFFICER
ARRESTING OFFICER AGENCY: North Port PD
LAST AGENCY INSPECTION: 05/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 08:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:00

SUBJECT'S LAST NAME: GARRAUS

DATE/TIME OF TEST: 06/10/2017 10:31:57

Diagnostics Check		OK	10:25
Air Blank		0.000	10:26
Control Test		0.081	10:26
Air Blank		0.000	10:26
Subject Sample	# 1	0.215	10:27
Breath Volume		3.875	10:27
Air Blank		0.000	10:28
Air Blank		0.000	10:29
Subject Sample	# 2	0.215	10:30
Breath Volume		3.394	10:30
Air Blank		0.000	10:31
Control Test		0.079	10:31
Air Blank		0.000	10:31
Diagnostics Check		OK	10:31

Data Download Date/Time: 06/28/2017 07:30

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,DANIEL,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: REESE,DEPUTY
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 05/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:47

SUBJECT'S LAST NAME: HURLEY

DATE/TIME OF TEST: 06/11/2017 04:44:24

Diagnostics Check		OK	04:33
Air Blank		0.000	04:34
Control Test		0.079	04:34
Air Blank		0.000	04:35
Subject Sample	# 1	0.210	04:38
Breath Volume		1.984	04:38
Air Blank		0.000	04:39
Air Blank		0.000	04:40
Subject Sample	# 2	0.213	04:42
Breath Volume		1.175	04:42
Air Blank		0.000	04:43
Control Test		0.077	04:43
Air Blank		0.000	04:44
Diagnostics Check		OK	04:44

Data Download Date/Time: 06/28/2017 07:30

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SLOCUM,RANDAL,L
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: SLOCUM,RANDAL
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 05/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:10

SUBJECT'S LAST NAME: RACHUBA

DATE/TIME OF TEST: 06/11/2017 21:57:13

Diagnostics Check		OK	21:49
Air Blank		0.000	21:50
Control Test		0.080	21:50
Air Blank		0.000	21:51
Subject Sample	# 1	0.277	21:52
Breath Volume		3.132	21:52
Air Blank		0.000	21:53
Air Blank		0.000	21:55
Subject Sample	# 2	0.279	21:55
Breath Volume		2.996	21:55
Air Blank		0.000	21:56
Control Test		0.077	21:56
Air Blank		0.000	21:57
Diagnostics Check		OK	21:57

Data Download Date/Time: 06/28/2017 07:30

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZCOLON,DANIEL,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BJORK,OFFICER
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 05/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:37

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 06/11/2017 23:15:36

Diagnostics Check		OK	23:09
Air Blank		0.000	23:09
Control Test		0.080	23:10
Air Blank		0.000	23:10
Subject Sample	# 1	0.389	23:10
Breath Volume		2.382	23:10
Air Blank		0.000	23:11
Air Blank		0.000	23:13
Subject Sample	# 2	0.393	23:13
Breath Volume		3.308	23:13
Air Blank		0.000	23:14
Control Test		0.079	23:15
Air Blank		0.000	23:15
Diagnostics Check		OK	23:15

Data Download Date/Time: 06/28/2017 07:30

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BABCOCK,TAMRA,J
BREATH TEST OPERATOR AGENCY: North Port PD
ARRESTING OFFICER: BOETGER,ELI
ARRESTING OFFICER AGENCY: North Port PD
LAST AGENCY INSPECTION: 05/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: BABCOCK

DATE/TIME OF TEST: 06/14/2017 03:29:38

Diagnostics Check		OK	03:18	
Air Blank		0.000	03:18	
Control Test		0.080	03:18	
Air Blank		0.000	03:19	
Subject Sample	# 1	0.141	03:22	Volume Not Met
Breath Volume		0.148	03:22	
Air Blank		0.000	03:23	
Air Blank		0.000	03:24	
Subject Sample	# 2		03:28	No Sample Provided
Breath Volume		0.000	03:28	
Air Blank		0.000	03:28	
Control Test		0.078	03:29	
Air Blank		0.000	03:29	
Diagnostics Check		OK	03:29	

Data Download Date/Time: 06/28/2017 07:31

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,JIMMY,J
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: ADAMS,JIMMY
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 05/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 09:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:21

SUBJECT'S LAST NAME: FLEMING

DATE/TIME OF TEST: 06/14/2017 10:48:34

Diagnostics Check		OK	10:42
Air Blank		0.000	10:43
Control Test		0.079	10:43
Air Blank		0.000	10:44
Subject Sample	# 1	0.000	10:44
Breath Volume		1.492	10:44
Air Blank		0.000	10:44
Air Blank		0.000	10:46
Subject Sample	# 2	0.000	10:47
Breath Volume		1.265	10:47
Air Blank		0.000	10:47
Control Test		0.079	10:47
Air Blank		0.000	10:48
Diagnostics Check		OK	10:48

Data Download Date/Time: 06/28/2017 07:31

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PETERSON,STEFANIE,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: PETERSON,STEFANIE
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 05/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 16:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:47

SUBJECT'S LAST NAME: CHOPKI

DATE/TIME OF TEST: 06/16/2017 18:14:07

Diagnostics Check		OK	18:07
Air Blank		0.000	18:08
Control Test		0.081	18:08
Air Blank		0.000	18:09
Subject Sample	# 1	0.164	18:09
Breath Volume		3.578	18:09
Air Blank		0.000	18:10
Air Blank		0.000	18:11
Subject Sample	# 2	0.165	18:12
Breath Volume		3.476	18:12
Air Blank		0.000	18:13
Control Test		0.080	18:13
Air Blank		0.000	18:14
Diagnostics Check		OK	18:14

Data Download Date/Time: 06/28/2017 07:31

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PETERSON,STEFANIE,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: PETERSON,STEFANIE
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 05/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 16:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:08

SUBJECT'S LAST NAME: SHAW

DATE/TIME OF TEST: 06/17/2017 17:35:00

Diagnostics Check		OK	17:28
Air Blank		0.000	17:29
Control Test		0.081	17:29
Air Blank		0.000	17:30
Subject Sample	# 1	0.204	17:30
Breath Volume		2.222	17:30
Air Blank		0.000	17:31
Air Blank		0.000	17:32
Subject Sample	# 2	0.204	17:33
Breath Volume		1.609	17:33
Air Blank		0.000	17:34
Control Test		0.081	17:34
Air Blank		0.000	17:34
Diagnostics Check		OK	17:35

Data Download Date/Time: 06/28/2017 07:31

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURLEY,SHAUN,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: MARTIN,SHAUN
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 05/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 06/18/2017 04:08:14

Diagnostics Check		OK	03:56	
Air Blank		0.000	03:57	
Control Test		0.080	03:57	
Air Blank		0.000	03:58	
Subject Sample	# 1		04:01	No Sample Provided
Breath Volume		0.000	04:01	
Air Blank		0.000	04:01	
Air Blank		0.000	04:03	
Subject Sample	# 2		04:06	No Sample Provided
Breath Volume		0.000	04:06	
Air Blank		0.000	04:07	
Control Test		0.081	04:07	
Air Blank		0.000	04:08	
Diagnostics Check		OK	04:08	

Data Download Date/Time: 06/28/2017 07:31

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NAMORS,JASON,R
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: HYDE,E
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 05/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:14

SUBJECT'S LAST NAME: CURRY

DATE/TIME OF TEST: 06/18/2017 23:54:21

Diagnostics Check		OK	23:48
Air Blank		0.000	23:48
Control Test		0.080	23:48
Air Blank		0.000	23:49
Subject Sample	# 1	0.175	23:49
Breath Volume		1.496	23:49
Air Blank		0.000	23:50
Air Blank		0.000	23:52
Subject Sample	# 2	0.191	23:52
Breath Volume		2.488	23:52
Air Blank		0.000	23:53
Control Test		0.080	23:53
Air Blank		0.000	23:54
Diagnostics Check		OK	23:54

Data Download Date/Time: 06/28/2017 07:31

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANIK,JONATHAN,F
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: VANIK,JONATHAN
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 05/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:12

SUBJECT'S LAST NAME: BOYETT

DATE/TIME OF TEST: 06/19/2017 01:46:21

Diagnostics Check		OK	01:39
Air Blank		0.000	01:40
Control Test		0.079	01:40
Air Blank		0.000	01:41
Subject Sample	# 1	0.149	01:41
Breath Volume		3.128	01:41
Air Blank		0.000	01:42
Air Blank		0.000	01:44
Subject Sample	# 2	0.150	01:44
Breath Volume		3.507	01:44
Air Blank		0.000	01:45
Control Test		0.079	01:45
Air Blank		0.000	01:46
Diagnostics Check		OK	01:46

Data Download Date/Time: 06/28/2017 07:31

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 05/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: POLOA

DATE/TIME OF TEST: 06/20/2017 02:59:33

Diagnostics Check		OK	02:52
Air Blank		0.000	02:52
Control Test		0.081	02:52
Air Blank		0.000	02:53
Subject Sample	# 1	0.212	02:54
Breath Volume		1.386	02:54
Air Blank		0.000	02:55
Air Blank		0.000	02:56
Subject Sample	# 2	0.225	02:57
Breath Volume		2.464	02:57
Air Blank		0.000	02:58
Control Test		0.077	02:58
Air Blank		0.000	02:59
Diagnostics Check		OK	02:59

Data Download Date/Time: 06/28/2017 07:31

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HEYMAN,LUKE,S
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: HEYMAN,LUKE
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 05/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 09:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:50

SUBJECT'S LAST NAME: SHELEY

DATE/TIME OF TEST: 06/21/2017 10:21:22

Diagnostics Check		OK	10:15
Air Blank		0.000	10:15
Control Test		0.080	10:15
Air Blank		0.000	10:16
Subject Sample	# 1	0.000	10:17
Breath Volume		4.355	10:17
Air Blank		0.000	10:17
Air Blank		0.000	10:19
Subject Sample	# 2	0.000	10:20
Breath Volume		3.636	10:20
Air Blank		0.000	10:20
Control Test		0.079	10:20
Air Blank		0.000	10:21
Diagnostics Check		OK	10:21

Data Download Date/Time: 06/28/2017 07:31

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURLEY,SHAUN,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BUEHLER,M
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 05/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:51

SUBJECT'S LAST NAME: BURNS

DATE/TIME OF TEST: 06/28/2017 02:22:16

Diagnostics Check		OK	02:14
Air Blank		0.000	02:15
Control Test		0.080	02:15
Air Blank		0.000	02:15
Subject Sample	# 1	0.117	02:17
Breath Volume		1.601	02:17
Air Blank		0.000	02:18
Air Blank		0.000	02:20
Subject Sample	# 2	0.117	02:20
Breath Volume		2.148	02:20
Air Blank		0.000	02:21
Control Test		0.078	02:21
Air Blank		0.000	02:22
Diagnostics Check		OK	02:22

Data Download Date/Time: 06/28/2017 07:31

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUBRANO,DEP,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: SLOWICK,C
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 06/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:42

SUBJECT'S LAST NAME: DE VITA

DATE/TIME OF TEST: 06/28/2017 21:12:49

Diagnostics Check		OK	21:06
Air Blank		0.000	21:07
Control Test		0.080	21:07
Air Blank		0.000	21:08
Subject Sample	# 1	0.000	21:08
Breath Volume		1.746	21:08
Air Blank		0.000	21:09
Air Blank		0.000	21:11
Subject Sample	# 2	0.000	21:11
Breath Volume		1.367	21:11
Air Blank		0.000	21:11
Control Test		0.081	21:12
Air Blank		0.000	21:12
Diagnostics Check		OK	21:12

Data Download Date/Time: 07/28/2017 07:56

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,NICKKI,N
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BUEHLER,M
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 06/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:56

SUBJECT'S LAST NAME: GRANDE

DATE/TIME OF TEST: 06/29/2017 00:33:01

Diagnostics Check		OK	00:25
Air Blank		0.000	00:26
Control Test		0.080	00:26
Air Blank		0.000	00:26
Subject Sample	# 1	0.178	00:28
Breath Volume		1.707	00:28
Air Blank		0.000	00:28
Air Blank		0.000	00:30
Subject Sample	# 2	0.182	00:31
Breath Volume		1.441	00:31
Air Blank		0.000	00:32
Control Test		0.079	00:32
Air Blank		0.000	00:32
Diagnostics Check		OK	00:33

Data Download Date/Time: 07/28/2017 07:56

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,LEDFORD
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 06/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: PARK

DATE/TIME OF TEST: 06/29/2017 00:43:35

Diagnostics Check		OK	00:40	
Air Blank		0.000	00:41	
Control Test		0.079	00:41	
Air Blank		0.000	00:41	
Subject Sample	# 1		00:42	Subject Test Refused
Breath Volume		0.000	00:42	
Air Blank		0.000	00:42	
Control Test		0.080	00:42	
Air Blank		0.000	00:43	
Diagnostics Check		OK	00:43	

Data Download Date/Time: 07/28/2017 07:56

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NAMORS,JASON,R
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: SCHUMANN,D
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 06/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:02

SUBJECT'S LAST NAME: GOLEMBESKI

DATE/TIME OF TEST: 06/30/2017 03:39:11

Diagnostics Check		OK	03:31
Air Blank		0.000	03:32
Control Test		0.079	03:32
Air Blank		0.000	03:33
Subject Sample	# 1	0.189	03:33
Breath Volume		2.457	03:33
Air Blank		0.000	03:34
Air Blank		0.000	03:36
Subject Sample	# 2	0.192	03:37
Breath Volume		2.476	03:37
Air Blank		0.000	03:38
Control Test		0.077	03:38
Air Blank		0.000	03:39
Diagnostics Check		OK	03:39

Data Download Date/Time: 07/28/2017 07:56

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURLEY,SHAUN,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: OCTAVIAN,K
ARRESTING OFFICER AGENCY: North Port PD
LAST AGENCY INSPECTION: 06/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: DAY

DATE/TIME OF TEST: 06/30/2017 23:13:40

Diagnostics Check		OK	23:06
Air Blank		0.000	23:06
Control Test		0.081	23:06
Air Blank		0.000	23:07
Subject Sample	# 1	0.228	23:08
Breath Volume		1.816	23:08
Air Blank		0.000	23:09
Air Blank		0.000	23:11
Subject Sample	# 2	0.219	23:12
Breath Volume		2.320	23:12
Air Blank		0.000	23:12
Control Test		0.080	23:13
Air Blank		0.000	23:13
Diagnostics Check		OK	23:13

Data Download Date/Time: 07/28/2017 07:56

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

INSTRUMENT REGISTERED TO: EUSTIS P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRIEVES,MATTHEW,J
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: LEDLOW,AARON
ARRESTING OFFICER AGENCY: Eustis PD
LAST AGENCY INSPECTION: 05/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 22:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:14

SUBJECT'S LAST NAME: WASHO

DATE/TIME OF TEST: 06/03/2017 23:40:15

Diagnostics Check		OK	23:36	
Air Blank		0.000	23:36	
Control Test		0.079	23:36	
Air Blank		0.000	23:37	
Subject Sample	# 1		23:38	Subject Test Refused
Breath Volume		0.000	23:38	
Air Blank		0.000	23:39	
Control Test		0.080	23:39	
Air Blank		0.000	23:40	
Diagnostics Check		OK	23:40	

Data Download Date/Time: 06/26/2017 10:01

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

INSTRUMENT REGISTERED TO: EUSTIS P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HORST,CHRIS,A
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: ARNOLD,LEONARD
ARRESTING OFFICER AGENCY: Eustis PD
LAST AGENCY INSPECTION: 06/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 01:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:14

SUBJECT'S LAST NAME: MCKINNEY

DATE/TIME OF TEST: 06/29/2017 01:41:23

Diagnostics Check		OK	01:35
Air Blank		0.000	01:35
Control Test		0.079	01:36
Air Blank		0.000	01:36
Subject Sample	# 1	0.155	01:37
Breath Volume		1.589	01:37
Air Blank		0.000	01:37
Air Blank		0.000	01:39
Subject Sample	# 2	0.166	01:39
Breath Volume		2.679	01:39
Air Blank		0.000	01:40
Control Test		0.078	01:40
Air Blank		0.000	01:41
Diagnostics Check		OK	01:41

Data Download Date/Time: 07/24/2017 09:40

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSTYAL,MATT,A
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: KOSTYAL,MATT
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 05/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: MILLER JR

DATE/TIME OF TEST: 06/04/2017 01:18:32

Diagnostics Check		OK	01:12
Air Blank		0.000	01:12
Control Test		0.078	01:13
Air Blank		0.000	01:13
Subject Sample	# 1	0.000	01:14
Breath Volume		3.746	01:14
Air Blank		0.000	01:14
Air Blank		0.000	01:16
Subject Sample	# 2	0.000	01:17
Breath Volume		3.136	01:17
Air Blank		0.000	01:17
Control Test		0.078	01:17
Air Blank		0.000	01:18
Diagnostics Check		OK	01:18

Data Download Date/Time: 06/28/2017 08:04

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSTYAL,MATT,A
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: KOSTYAL,MATT
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 05/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: OASTER

DATE/TIME OF TEST: 06/06/2017 22:56:16

Diagnostics Check		OK	22:50
Air Blank		0.000	22:50
Control Test		0.078	22:51
Air Blank		0.000	22:51
Subject Sample	# 1	0.159	22:51
Breath Volume		3.195	22:51
Air Blank		0.000	22:52
Air Blank		0.000	22:54
Subject Sample	# 2	0.162	22:54
Breath Volume		3.113	22:54
Air Blank		0.000	22:55
Control Test		0.078	22:55
Air Blank		0.000	22:56
Diagnostics Check		OK	22:56

Data Download Date/Time: 06/28/2017 08:04

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSTYAL,MATT,A
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: KOSTYAL,MATT
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 05/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:28

SUBJECT'S LAST NAME: CLARK

DATE/TIME OF TEST: 06/08/2017 20:54:34

Diagnostics Check		OK	20:48
Air Blank		0.000	20:48
Control Test		0.079	20:49
Air Blank		0.000	20:49
Subject Sample	# 1	0.152	20:50
Breath Volume		3.621	20:50
Air Blank		0.000	20:50
Air Blank		0.000	20:52
Subject Sample	# 2	0.153	20:53
Breath Volume		3.742	20:53
Air Blank		0.000	20:53
Control Test		0.079	20:53
Air Blank		0.000	20:54
Diagnostics Check		OK	20:54

Data Download Date/Time: 06/28/2017 08:04

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLINT,SERGEANT,C
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: KOSTYAL,DEPUTY
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 05/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:47

SUBJECT'S LAST NAME: PORTEE

DATE/TIME OF TEST: 06/09/2017 00:19:05

Diagnostics Check		OK	00:11
Air Blank		0.000	00:11
Control Test		0.078	00:12
Air Blank		0.000	00:12
Subject Sample	# 1	0.097	00:14
Breath Volume		1.761	00:14
Air Blank		0.000	00:14
Air Blank		0.000	00:16
Subject Sample	# 2	0.095	00:17
Breath Volume		1.507	00:17
Air Blank		0.000	00:18
Control Test		0.077	00:18
Air Blank		0.000	00:19
Diagnostics Check		OK	00:19

Data Download Date/Time: 06/28/2017 08:04

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSTYAL,MATT,A
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: KOSTYAL,MATT
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 05/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:52

SUBJECT'S LAST NAME: CALDWELL

DATE/TIME OF TEST: 06/20/2017 20:20:22

Diagnostics Check		OK	20:12	
Air Blank		0.000	20:13	
Control Test		0.077	20:13	
Air Blank		0.000	20:14	
Subject Sample	# 1	0.245	20:15	
Breath Volume		1.761	20:15	
Air Blank		0.000	20:16	
Air Blank		0.000	20:18	
Subject Sample	# 2		20:18	Slope Not Met
Breath Volume		0.960	20:18	
Air Blank		0.000	20:19	
Control Test		0.077	20:19	
Air Blank		0.000	20:20	
Diagnostics Check		OK	20:20	

Data Download Date/Time: 06/28/2017 08:04

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSTYAL,MATT,A
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: KOSTYAL,MATT
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 05/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:52

SUBJECT'S LAST NAME: CALDWELL

DATE/TIME OF TEST: 06/20/2017 20:33:37

Diagnostics Check		OK	20:26
Air Blank		0.000	20:27
Control Test		0.078	20:27
Air Blank		0.000	20:28
Subject Sample	# 1	0.260	20:28
Breath Volume		3.660	20:28
Air Blank		0.000	20:29
Air Blank		0.000	20:31
Subject Sample	# 2	0.255	20:32
Breath Volume		1.308	20:32
Air Blank		0.000	20:32
Control Test		0.077	20:33
Air Blank		0.000	20:33
Diagnostics Check		OK	20:33

Data Download Date/Time: 06/28/2017 08:04

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001361

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HITCHCOCK,WILLIAM,D
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: HITCHCOCK,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 06/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 539031
DRY GAS EXPIRY DATE: 08/25/2019
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: ANDRESS

DATE/TIME OF TEST: 06/28/2017 00:28:35

Diagnostics Check		OK	00:21
Air Blank		0.000	00:22
Control Test		0.081	00:22
Air Blank		0.000	00:22
Subject Sample	# 1	0.208	00:23
Breath Volume		1.699	00:23
Air Blank		0.000	00:24
Air Blank		0.000	00:26
Subject Sample	# 2	0.197	00:27
Breath Volume		1.308	00:27
Air Blank		0.000	00:27
Control Test		0.081	00:27
Air Blank		0.000	00:28
Diagnostics Check		OK	00:28

Data Download Date/Time: 07/31/2017 14:31

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001361

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HITCHCOCK,WILLIAM,D
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: HITCHCOCK,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 06/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 539031
DRY GAS EXPIRY DATE: 08/25/2019
VIOLATION CODE: DUI
ARREST TIME: 00:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: MCCURLEY

DATE/TIME OF TEST: 06/30/2017 01:02:44

Diagnostics Check		OK	00:53	
Air Blank		0.000	00:53	
Control Test		0.081	00:54	
Air Blank		0.000	00:54	
Subject Sample	# 1		00:57	No Sample Provided
Breath Volume		0.000	00:57	
Air Blank		0.000	00:58	
Air Blank		0.000	01:00	
Subject Sample	# 2		01:01	Subject Test Refused
Breath Volume		0.000	01:01	
Air Blank		0.000	01:01	
Control Test		0.080	01:02	
Air Blank		0.000	01:02	
Diagnostics Check		OK	01:02	

Data Download Date/Time: 07/31/2017 14:31

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARKER,JEREMY,A
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: FARLEY,OFFICER
ARRESTING OFFICER AGENCY: Punta Gorda PD
LAST AGENCY INSPECTION: 05/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A6
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:59

SUBJECT'S LAST NAME: TROXEL

DATE/TIME OF TEST: 06/03/2017 03:26:24

Diagnostics Check		OK	03:19
Air Blank		0.000	03:20
Control Test		0.080	03:20
Air Blank		0.000	03:21
Subject Sample	# 1	0.100	03:21
Breath Volume		1.300	03:21
Air Blank		0.000	03:22
Air Blank		0.000	03:24
Subject Sample	# 2	0.106	03:24
Breath Volume		2.054	03:24
Air Blank		0.000	03:25
Control Test		0.080	03:25
Air Blank		0.000	03:26
Diagnostics Check		OK	03:26

Data Download Date/Time: 06/14/2017 11:02

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KLOS,LINDSAY,M
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: MCGLASHAN,DEPUTY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 05/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A6
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:55

SUBJECT'S LAST NAME: ROBERSON

DATE/TIME OF TEST: 06/03/2017 21:22:05

Diagnostics Check		OK	21:15
Air Blank		0.000	21:15
Control Test		0.079	21:15
Air Blank		0.000	21:16
Subject Sample	# 1	0.174	21:17
Breath Volume		1.296	21:17
Air Blank		0.000	21:17
Air Blank		0.000	21:19
Subject Sample	# 2	0.192	21:20
Breath Volume		2.593	21:20
Air Blank		0.000	21:21
Control Test		0.079	21:21
Air Blank		0.000	21:22
Diagnostics Check		OK	21:22

Data Download Date/Time: 06/14/2017 11:02

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARKER,JEREMY,A
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: FARLEY,OFFICER
ARRESTING OFFICER AGENCY: Punta Gorda PD
LAST AGENCY INSPECTION: 05/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A6
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:39

SUBJECT'S LAST NAME: BUNKE

DATE/TIME OF TEST: 06/03/2017 23:06:09

Diagnostics Check		OK	22:59
Air Blank		0.000	23:00
Control Test		0.080	23:00
Air Blank		0.000	23:01
Subject Sample	# 1	0.155	23:01
Breath Volume		2.144	23:01
Air Blank		0.000	23:02
Air Blank		0.000	23:04
Subject Sample	# 2	0.155	23:04
Breath Volume		3.074	23:04
Air Blank		0.000	23:05
Control Test		0.080	23:05
Air Blank		0.000	23:06
Diagnostics Check		OK	23:06

Data Download Date/Time: 06/14/2017 11:02

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARKER,JEREMY,A
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: FARLEY,OFFICER
ARRESTING OFFICER AGENCY: Punta Gorda PD
LAST AGENCY INSPECTION: 05/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A6
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: WEST

DATE/TIME OF TEST: 06/04/2017 03:48:27

Diagnostics Check		OK	03:36	
Air Blank		0.000	03:36	
Control Test		0.080	03:37	
Air Blank		0.000	03:37	
Subject Sample	# 1	0.058	03:38	
Breath Volume		1.574	03:38	
Air Blank		0.000	03:38	
Air Blank		0.000	03:40	
Subject Sample	# 2	0.053	03:44	Volume Not Met
Breath Volume		0.234	03:44	
Air Blank		0.000	03:44	
Air Blank		0.000	03:46	
Subject Sample	# 3	0.054	03:46	
Breath Volume		1.859	03:46	
Air Blank		0.000	03:47	
Control Test		0.078	03:47	
Air Blank		0.000	03:48	
Diagnostics Check		OK	03:48	

Data Download Date/Time: 06/14/2017 11:02

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PATTON,JOHN,G
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: WRASSE,NICHOLAS
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 05/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A6
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:42

SUBJECT'S LAST NAME: KOWALEWSKI

DATE/TIME OF TEST: 06/04/2017 04:09:31

Diagnostics Check		OK	04:02
Air Blank		0.000	04:03
Control Test		0.079	04:03
Air Blank		0.000	04:04
Subject Sample	# 1	0.245	04:04
Breath Volume		1.718	04:04
Air Blank		0.000	04:05
Air Blank		0.000	04:07
Subject Sample	# 2	0.246	04:08
Breath Volume		1.195	04:08
Air Blank		0.000	04:08
Control Test		0.079	04:08
Air Blank		0.000	04:09
Diagnostics Check		OK	04:09

Data Download Date/Time: 06/14/2017 11:02

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCLAIN,CHRISTOPHER,M
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: HARTMANSTORFER,C
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 05/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A6
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:36

SUBJECT'S LAST NAME: RHEAUME

DATE/TIME OF TEST: 06/06/2017 02:11:07

Diagnostics Check		OK	01:58	
Air Blank		0.000	01:58	
Control Test		0.080	01:59	
Air Blank		0.000	01:59	
Subject Sample	# 1	0.135	02:00	
Breath Volume		1.218	02:00	
Air Blank		0.000	02:01	
Air Blank		0.000	02:03	
Subject Sample	# 2	0.133	02:06	Volume Not Met
Breath Volume		0.613	02:06	
Air Blank		0.000	02:07	
Air Blank		0.000	02:09	
Subject Sample	# 3	0.147	02:09	
Breath Volume		2.765	02:09	
Air Blank		0.000	02:10	
Control Test		0.079	02:10	
Air Blank		0.000	02:11	
Diagnostics Check		OK	02:11	

Data Download Date/Time: 06/14/2017 11:02

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMERON,CELESTE,M
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: LIBBY,JEROME
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 05/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A6
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 12:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:00

SUBJECT'S LAST NAME: MAJOR

DATE/TIME OF TEST: 06/06/2017 13:27:52

Diagnostics Check	OK	13:27	
Air Blank		13:27	RFI Detect
Air Blank	0.000	13:27	

Data Download Date/Time: 06/14/2017 11:02

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMERON,CELESTE,M
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: LIBBY,JEROME
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 05/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A6
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 12:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:00

SUBJECT'S LAST NAME: MAJOR

DATE/TIME OF TEST: 06/06/2017 13:30:36

Diagnostics Check	OK	13:29	
Air Blank		13:30	RFI Detect
Air Blank		13:30	RFI Detect

Data Download Date/Time: 06/14/2017 11:02

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMERON,CELESTE,M
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: LIBBY,JEROME
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 05/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A6
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 12:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:00

SUBJECT'S LAST NAME: MAJOR

DATE/TIME OF TEST: 06/06/2017 13:45:02

Diagnostics Check		OK	13:39
Air Blank		0.000	13:39
Control Test		0.079	13:39
Air Blank		0.000	13:40
Subject Sample	# 1	0.288	13:40
Breath Volume		1.273	13:40
Air Blank		0.000	13:41
Air Blank		0.000	13:43
Subject Sample	# 2	0.269	13:43
Breath Volume		1.308	13:43
Air Blank		0.000	13:44
Control Test		0.080	13:44
Air Blank		0.000	13:44
Diagnostics Check		OK	13:45

Data Download Date/Time: 06/14/2017 11:02

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EGAN,MELISSA,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: HARTMANSTORFER,CHRIS
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 05/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A6
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:21

SUBJECT'S LAST NAME: DAHL

DATE/TIME OF TEST: 06/06/2017 22:57:58

Diagnostics Check		OK	22:51
Air Blank		0.000	22:51
Control Test		0.080	22:52
Air Blank		0.000	22:52
Subject Sample	# 1	0.177	22:52
Breath Volume		1.722	22:52
Air Blank		0.000	22:53
Air Blank		0.000	22:55
Subject Sample	# 2	0.178	22:56
Breath Volume		1.921	22:56
Air Blank		0.000	22:57
Control Test		0.079	22:57
Air Blank		0.000	22:57
Diagnostics Check		OK	22:57

Data Download Date/Time: 06/14/2017 11:02

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMBRIA,STEVEN,J
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: KATSARELAS,K
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 05/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A6
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:25

SUBJECT'S LAST NAME: GAST

DATE/TIME OF TEST: 06/07/2017 21:56:23

Diagnostics Check		OK	21:48
Air Blank		0.000	21:49
Control Test		0.080	21:49
Air Blank		0.000	21:50
Subject Sample	# 1	0.109	21:51
Breath Volume		2.109	21:51
Air Blank		0.000	21:52
Air Blank		0.000	21:54
Subject Sample	# 2	0.103	21:54
Breath Volume		1.566	21:54
Air Blank		0.000	21:55
Control Test		0.079	21:55
Air Blank		0.000	21:56
Diagnostics Check		OK	21:56

Data Download Date/Time: 06/14/2017 11:02

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARKER,JEREMY,A
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: WRASSE,OFFICER
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 05/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A6
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: BORG

DATE/TIME OF TEST: 06/09/2017 01:43:02

Diagnostics Check		OK	01:35
Air Blank		0.000	01:35
Control Test		0.080	01:36
Air Blank		0.000	01:36
Subject Sample	# 1	0.238	01:37
Breath Volume		2.433	01:37
Air Blank		0.000	01:37
Air Blank		0.000	01:39
Subject Sample	# 2	0.242	01:41
Breath Volume		2.667	01:41
Air Blank		0.000	01:42
Control Test		0.079	01:42
Air Blank		0.000	01:42
Diagnostics Check		OK	01:43

Data Download Date/Time: 06/14/2017 11:02

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCLAIN,CHRISTOPHER,M
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: MORAGHAN,J
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 05/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A6
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 18:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:46

SUBJECT'S LAST NAME: CASCIELLO

DATE/TIME OF TEST: 06/09/2017 19:15:51

Diagnostics Check		OK	19:08
Air Blank		0.000	19:09
Control Test		0.079	19:09
Air Blank		0.000	19:10
Subject Sample	# 1	0.253	19:11
Breath Volume		2.511	19:11
Air Blank		0.000	19:11
Air Blank		0.000	19:13
Subject Sample	# 2	0.246	19:14
Breath Volume		2.148	19:14
Air Blank		0.000	19:14
Control Test		0.078	19:15
Air Blank		0.000	19:15
Diagnostics Check		OK	19:15

Data Download Date/Time: 06/14/2017 11:02

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCLAIN,CHRISTOPHER,M
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: HARTMANSTOFER,C
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 05/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A6
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:07

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 06/09/2017 23:40:01

Diagnostics Check		OK	23:28	
Air Blank		0.000	23:28	
Control Test		0.080	23:29	
Air Blank		0.000	23:29	
Subject Sample	# 1	0.162	23:32	Volume Not Met
Breath Volume		0.015	23:32	
Air Blank		0.000	23:33	
Air Blank		0.000	23:35	
Subject Sample	# 2	0.152	23:38	Volume Not Met
Breath Volume		0.015	23:38	
Air Blank		0.000	23:39	
Control Test		0.078	23:39	
Air Blank		0.000	23:39	
Diagnostics Check		OK	23:40	

Data Download Date/Time: 06/14/2017 11:02

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCLAIN,CHRISTOPHER,M
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: MORAGHAN,J
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 05/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A6
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:14

SUBJECT'S LAST NAME: WELZ

DATE/TIME OF TEST: 06/10/2017 02:49:23

Diagnostics Check		OK	02:35	
Air Blank		0.000	02:36	
Control Test		0.079	02:36	
Air Blank		0.000	02:37	
Subject Sample	# 1	0.069	02:37	
Breath Volume		1.632	02:37	
Air Blank		0.000	02:38	
Air Blank		0.000	02:40	
Subject Sample	# 2	0.058	02:43	Volume Not Met
Breath Volume		0.636	02:43	
Air Blank		0.000	02:43	
Air Blank		0.000	02:45	
Subject Sample	# 3	0.064	02:47	
Breath Volume		1.820	02:47	
Air Blank		0.000	02:48	
Control Test		0.079	02:48	
Air Blank		0.000	02:49	
Diagnostics Check		OK	02:49	

Data Download Date/Time: 06/14/2017 11:02

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EGAN,MELISSA,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: ADAMCHIK,L
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 05/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A6
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:22

SUBJECT'S LAST NAME: BETTS

DATE/TIME OF TEST: 06/10/2017 04:54:33

Diagnostics Check		OK	04:48
Air Blank		0.000	04:48
Control Test		0.079	04:48
Air Blank		0.000	04:49
Subject Sample	# 1	0.158	04:49
Breath Volume		1.800	04:49
Air Blank		0.000	04:50
Air Blank		0.000	04:52
Subject Sample	# 2	0.151	04:53
Breath Volume		1.621	04:53
Air Blank		0.000	04:53
Control Test		0.079	04:53
Air Blank		0.000	04:54
Diagnostics Check		OK	04:54

Data Download Date/Time: 06/14/2017 11:02

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMYRNIOS,MANUEL,J
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: SMYRNIOS,MANUEL
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 05/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A6
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:25

SUBJECT'S LAST NAME: WYRICK

DATE/TIME OF TEST: 06/10/2017 05:53:06

Diagnostics Check		OK	05:45	
Air Blank		0.000	05:46	
Control Test		0.079	05:46	
Air Blank		0.000	05:47	
Subject Sample	# 1		05:50	No Sample Provided
Breath Volume		0.000	05:50	
Air Blank		0.000	05:50	
Air Blank			05:52	RFI Detect
Air Blank			05:52	RFI Detect
Air Blank			05:53	RFI Detect

Data Download Date/Time: 06/14/2017 11:02

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHAPMAN,VANESSA,N
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: HARTMANSTORFER,CHRIS
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 05/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A6
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:36

SUBJECT'S LAST NAME: MOORE

DATE/TIME OF TEST: 06/12/2017 02:07:38

Diagnostics Check		OK	01:56	
Air Blank		0.000	01:56	
Control Test		0.080	01:57	
Air Blank		0.000	01:57	
Subject Sample	# 1	0.079	02:00	Volume Not Met
Breath Volume		0.031	02:00	
Air Blank		0.000	02:01	
Air Blank		0.000	02:03	
Subject Sample	# 2		02:06	No Sample Provided
Breath Volume		0.000	02:06	
Air Blank		0.000	02:06	
Control Test		0.079	02:07	
Air Blank		0.000	02:07	
Diagnostics Check		OK	02:07	

Data Download Date/Time: 06/14/2017 11:02

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCLAIN,CHRISTOPHER,M
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: PRESTON,S
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 05/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A6
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:36

SUBJECT'S LAST NAME: FLYNN

DATE/TIME OF TEST: 06/12/2017 04:59:04

Diagnostics Check	OK	04:57	
Air Blank	0.000	04:58	
Control Test	0.080	04:58	
Air Blank		04:58	RFI Detect
Air Blank	0.000	04:59	

Data Download Date/Time: 06/14/2017 11:02

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCLAIN,CHRISTOPHER,M
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: PRESTON,S
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 05/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A6
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:36

SUBJECT'S LAST NAME: FLYNN

DATE/TIME OF TEST: 06/12/2017 05:08:10

Diagnostics Check		OK	05:01
Air Blank		0.000	05:02
Control Test		0.080	05:02
Air Blank		0.000	05:02
Subject Sample	# 1	0.164	05:03
Breath Volume		4.179	05:03
Air Blank		0.000	05:04
Air Blank		0.000	05:06
Subject Sample	# 2	0.162	05:06
Breath Volume		4.007	05:06
Air Blank		0.000	05:07
Control Test		0.078	05:07
Air Blank		0.000	05:08
Diagnostics Check		OK	05:08

Data Download Date/Time: 06/14/2017 11:02

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCLAIN,CHRISTOPHER,M
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: MORAGHAN,J
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 06/14/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A6
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: LANDERS

DATE/TIME OF TEST: 06/14/2017 22:27:49

Diagnostics Check		OK	22:21
Air Blank		0.000	22:22
Control Test		0.080	22:22
Air Blank		0.000	22:22
Subject Sample	# 1	0.068	22:23
Breath Volume		4.460	22:23
Air Blank		0.000	22:23
Air Blank		0.000	22:25
Subject Sample	# 2	0.066	22:26
Breath Volume		2.089	22:26
Air Blank		0.000	22:26
Control Test		0.079	22:27
Air Blank		0.000	22:27
Diagnostics Check		OK	22:27

Data Download Date/Time: 07/25/2017 11:21

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRITES,CHRISTOPHER,J
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: SWARTZENTRUBER,BRANDON
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 06/14/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A6
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 15:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:29

SUBJECT'S LAST NAME: ORZECHOWSKI

DATE/TIME OF TEST: 06/15/2017 16:59:15

Diagnostics Check		OK	16:53
Air Blank		0.000	16:53
Control Test		0.080	16:53
Air Blank		0.000	16:54
Subject Sample	# 1	0.000	16:54
Breath Volume		1.914	16:54
Air Blank		0.000	16:55
Air Blank		0.000	16:57
Subject Sample	# 2	0.000	16:57
Breath Volume		2.281	16:57
Air Blank		0.000	16:58
Control Test		0.080	16:58
Air Blank		0.000	16:59
Diagnostics Check		OK	16:59

Data Download Date/Time: 07/25/2017 11:21

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHAPMAN,VANESSA,N
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: ADAMCHIK,LUKE
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 06/14/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A6
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: THOMA

DATE/TIME OF TEST: 06/16/2017 00:28:16

Diagnostics Check		OK	00:21
Air Blank		0.000	00:21
Control Test		0.080	00:22
Air Blank		0.000	00:22
Subject Sample	# 1	0.172	00:23
Breath Volume		1.839	00:23
Air Blank		0.000	00:24
Air Blank		0.000	00:26
Subject Sample	# 2	0.172	00:26
Breath Volume		1.500	00:26
Air Blank		0.000	00:27
Control Test		0.079	00:27
Air Blank		0.000	00:28
Diagnostics Check		OK	00:28

Data Download Date/Time: 07/25/2017 11:21

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FISHER,SCOTT,R
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: FARLEY,OFFICER
ARRESTING OFFICER AGENCY: Punta Gorda PD
LAST AGENCY INSPECTION: 06/14/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A6
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: PINTER

DATE/TIME OF TEST: 06/17/2017 01:09:11

Diagnostics Check		OK	01:01
Air Blank		0.000	01:01
Control Test		0.079	01:02
Air Blank		0.000	01:02
Subject Sample	# 1	0.097	01:04
Breath Volume		1.464	01:04
Air Blank		0.000	01:05
Air Blank		0.000	01:07
Subject Sample	# 2	0.099	01:07
Breath Volume		2.250	01:07
Air Blank		0.000	01:08
Control Test		0.079	01:08
Air Blank		0.000	01:09
Diagnostics Check		OK	01:09

Data Download Date/Time: 07/25/2017 11:21

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PATTON,JOHN,G
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: KATSARELAS,KEVIN
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 06/14/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A6
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: GREGGS

DATE/TIME OF TEST: 06/18/2017 01:07:17

Diagnostics Check		OK	00:55	
Air Blank		0.000	00:56	
Control Test		0.079	00:56	
Air Blank		0.000	00:57	
Subject Sample	# 1	0.170	01:00	Volume Not Met
Breath Volume		0.050	01:00	
Air Blank		0.000	01:00	
Air Blank		0.000	01:02	
Subject Sample	# 2	0.168	01:05	Volume Not Met
Breath Volume		0.050	01:05	
Air Blank		0.000	01:06	
Control Test		0.079	01:06	
Air Blank		0.000	01:07	
Diagnostics Check		OK	01:07	

Data Download Date/Time: 07/25/2017 11:21

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KLOS,LINDSAY,M
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: FARLEY,OFFICER
ARRESTING OFFICER AGENCY: Punta Gorda PD
LAST AGENCY INSPECTION: 06/14/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A6
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: FEINDEL

DATE/TIME OF TEST: 06/18/2017 02:45:42

Diagnostics Check		OK	02:35
Air Blank		0.000	02:36
Control Test		0.079	02:36
Air Blank		0.000	02:36
Subject Sample	# 1	0.229	02:39
Breath Volume		1.851	02:39
Air Blank		0.000	02:40
Air Blank		0.000	02:41
Subject Sample	# 2	0.231	02:44
Breath Volume		2.648	02:44
Air Blank		0.000	02:44
Control Test		0.078	02:45
Air Blank		0.000	02:45
Diagnostics Check		OK	02:45

Data Download Date/Time: 07/25/2017 11:21

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARRONE-CABRERA,ANTHONY,J
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: ROSE,GERALD
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 06/14/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A6
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 15:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:44

SUBJECT'S LAST NAME: PARISI

DATE/TIME OF TEST: 06/18/2017 17:15:50

Diagnostics Check		OK	17:09
Air Blank		0.000	17:09
Control Test		0.079	17:10
Air Blank		0.000	17:10
Subject Sample	# 1	0.172	17:11
Breath Volume		1.457	17:11
Air Blank		0.000	17:11
Air Blank		0.000	17:13
Subject Sample	# 2	0.169	17:14
Breath Volume		2.886	17:14
Air Blank		0.000	17:14
Control Test		0.078	17:15
Air Blank		0.000	17:15
Diagnostics Check		OK	17:15

Data Download Date/Time: 07/25/2017 11:21

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PLATZ,KEVIN,L
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: WEDGEWOOD,MATT
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 06/14/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A6
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 06/20/2017 21:33:19

Diagnostics Check		OK	21:26
Air Blank		0.000	21:26
Control Test		0.080	21:27
Air Blank		0.000	21:27
Subject Sample	# 1	0.058	21:28
Breath Volume		2.449	21:28
Air Blank		0.000	21:28
Air Blank		0.000	21:30
Subject Sample	# 2	0.058	21:31
Breath Volume		2.402	21:31
Air Blank		0.000	21:32
Control Test		0.080	21:32
Air Blank		0.000	21:33
Diagnostics Check		OK	21:33

Data Download Date/Time: 07/25/2017 11:21

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FISHER,SCOTT,R
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: MYRIE,DEPUTY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 06/14/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A6
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:37

SUBJECT'S LAST NAME: MARTYNENKO

DATE/TIME OF TEST: 06/22/2017 00:12:00

Diagnostics Check		OK	00:05
Air Blank		0.000	00:05
Control Test		0.080	00:05
Air Blank		0.000	00:06
Subject Sample	# 1	0.143	00:07
Breath Volume		1.578	00:07
Air Blank		0.000	00:08
Air Blank		0.000	00:10
Subject Sample	# 2	0.129	00:10
Breath Volume		1.582	00:10
Air Blank		0.000	00:11
Control Test		0.079	00:11
Air Blank		0.000	00:11
Diagnostics Check		OK	00:12

Data Download Date/Time: 07/25/2017 11:21

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCLAIN,CHRISTOPHER,M
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: KLEIVER,J
ARRESTING OFFICER AGENCY: Punta Gorda PD
LAST AGENCY INSPECTION: 06/14/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A6
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:27

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 06/26/2017 03:54:51

Diagnostics Check		OK	03:48
Air Blank		0.000	03:49
Control Test		0.080	03:49
Air Blank		0.000	03:49
Subject Sample	# 1	0.000	03:50
Breath Volume		1.531	03:50
Air Blank		0.000	03:50
Air Blank		0.000	03:52
Subject Sample	# 2	0.000	03:53
Breath Volume		1.355	03:53
Air Blank		0.000	03:53
Control Test		0.080	03:54
Air Blank		0.000	03:54
Diagnostics Check		OK	03:54

Data Download Date/Time: 07/25/2017 11:21

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EGAN,MELISSA,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: FARLEY,J
ARRESTING OFFICER AGENCY: Punta Gorda PD
LAST AGENCY INSPECTION: 06/14/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A6
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:20

SUBJECT'S LAST NAME: MANARINA

DATE/TIME OF TEST: 06/28/2017 20:54:09

Diagnostics Check		OK	20:47
Air Blank		0.000	20:48
Control Test		0.080	20:48
Air Blank		0.000	20:49
Subject Sample	# 1	0.184	20:49
Breath Volume		2.031	20:49
Air Blank		0.000	20:50
Air Blank		0.000	20:52
Subject Sample	# 2	0.186	20:52
Breath Volume		1.191	20:52
Air Blank		0.000	20:53
Control Test		0.079	20:53
Air Blank		0.000	20:54
Diagnostics Check		OK	20:54

Data Download Date/Time: 07/25/2017 11:21

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CREWS,HENRY,E
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: BARKER,THOMAS
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 06/14/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A6
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 11:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:55

SUBJECT'S LAST NAME: JANOVSKY

DATE/TIME OF TEST: 06/29/2017 12:23:18

Diagnostics Check		OK	12:17
Air Blank		0.000	12:17
Control Test		0.079	12:17
Air Blank		0.000	12:18
Subject Sample	# 1	0.000	12:19
Breath Volume		2.281	12:19
Air Blank		0.000	12:19
Air Blank		0.000	12:21
Subject Sample	# 2	0.000	12:21
Breath Volume		3.015	12:21
Air Blank		0.000	12:22
Control Test		0.080	12:22
Air Blank		0.000	12:23
Diagnostics Check		OK	12:23

Data Download Date/Time: 07/25/2017 11:21

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHAPMAN,VANESSA,N
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: REID,NICK
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 06/14/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A6
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 18:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:54

SUBJECT'S LAST NAME: RIVERS

DATE/TIME OF TEST: 06/29/2017 22:28:32

Diagnostics Check		OK	22:22
Air Blank		0.000	22:22
Control Test		0.080	22:23
Air Blank		0.000	22:23
Subject Sample	# 1	0.023	22:23
Breath Volume		1.531	22:23
Air Blank		0.000	22:24
Air Blank		0.000	22:26
Subject Sample	# 2	0.022	22:27
Breath Volume		1.371	22:27
Air Blank		0.000	22:27
Control Test		0.080	22:27
Air Blank		0.000	22:28
Diagnostics Check		OK	22:28

Data Download Date/Time: 07/25/2017 11:21

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCLAIN,CHRISTOPHER,M
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: FARLEY,J
ARRESTING OFFICER AGENCY: Punta Gorda PD
LAST AGENCY INSPECTION: 06/14/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A6
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: YDE

DATE/TIME OF TEST: 06/30/2017 03:39:04

Diagnostics Check		OK	03:32
Air Blank		0.000	03:32
Control Test		0.080	03:33
Air Blank		0.000	03:33
Subject Sample	# 1	0.010	03:34
Breath Volume		2.257	03:34
Air Blank		0.000	03:35
Air Blank		0.000	03:37
Subject Sample	# 2	0.010	03:37
Breath Volume		2.511	03:37
Air Blank		0.000	03:38
Control Test		0.081	03:38
Air Blank		0.000	03:38
Diagnostics Check		OK	03:39

Data Download Date/Time: 07/25/2017 11:21

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KLOS,LINDSAY,M
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: KATSARELAS,DEPUTY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 06/14/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A6
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: RILEY

DATE/TIME OF TEST: 06/30/2017 23:10:44

Diagnostics Check		OK	23:03
Air Blank		0.000	23:03
Control Test		0.080	23:03
Air Blank		0.000	23:04
Subject Sample	# 1	0.363	23:06
Breath Volume		2.054	23:06
Air Blank		0.000	23:06
Air Blank		0.000	23:08
Subject Sample	# 2	0.371	23:09
Breath Volume		1.921	23:09
Air Blank		0.000	23:09
Control Test		0.079	23:10
Air Blank		0.000	23:10
Diagnostics Check		OK	23:10

Data Download Date/Time: 07/25/2017 11:21

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEMEN,AARON,M
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: STEMEN,AARON
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 05/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: BERRESFORD

DATE/TIME OF TEST: 06/03/2017 03:26:33

Diagnostics Check		OK	03:14	
Air Blank		0.000	03:15	
Control Test		0.080	03:15	
Air Blank		0.000	03:16	
Subject Sample	# 1	0.113	03:19	Volume Not Met
Breath Volume		0.261	03:19	
Air Blank		0.000	03:20	
Air Blank		0.000	03:21	
Subject Sample	# 2	0.132	03:24	Volume Not Met
Breath Volume		0.101	03:24	
Air Blank		0.000	03:25	
Control Test		0.079	03:25	
Air Blank		0.000	03:26	
Diagnostics Check		OK	03:26	

Data Download Date/Time: 06/16/2017 20:06

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOSTICK,JESSE,W
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: SIMMONS,NAKEEMA
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 05/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:56

SUBJECT'S LAST NAME: GREEN

DATE/TIME OF TEST: 06/04/2017 02:22:57

Diagnostics Check		OK	02:16
Air Blank		0.000	02:16
Control Test		0.079	02:16
Air Blank		0.000	02:17
Subject Sample	# 1	0.172	02:17
Breath Volume		2.242	02:17
Air Blank		0.000	02:18
Air Blank		0.000	02:20
Subject Sample	# 2	0.166	02:21
Breath Volume		1.476	02:21
Air Blank		0.000	02:22
Control Test		0.078	02:22
Air Blank		0.000	02:22
Diagnostics Check		OK	02:22

Data Download Date/Time: 06/16/2017 20:06

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOSTICK,JESSE,W
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: MCFARLAND,COLLIN
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 05/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: WILKOWSKI

DATE/TIME OF TEST: 06/04/2017 02:57:12

Diagnostics Check		OK	02:51
Air Blank		0.000	02:51
Control Test		0.079	02:51
Air Blank		0.000	02:52
Subject Sample	# 1	0.206	02:52
Breath Volume		1.734	02:52
Air Blank		0.000	02:53
Air Blank		0.000	02:55
Subject Sample	# 2	0.201	02:55
Breath Volume		1.851	02:55
Air Blank		0.000	02:56
Control Test		0.079	02:56
Air Blank		0.000	02:57
Diagnostics Check		OK	02:57

Data Download Date/Time: 06/16/2017 20:06

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SOWARDS,JACOB,T
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: DYKES,HUDSON
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 05/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:45

SUBJECT'S LAST NAME: HOTT

DATE/TIME OF TEST: 06/07/2017 05:16:14

Diagnostics Check		OK	05:07
Air Blank		0.000	05:08
Control Test		0.080	05:08
Air Blank		0.000	05:09
Subject Sample	# 1	0.038	05:10
Breath Volume		1.292	05:10
Air Blank		0.000	05:11
Air Blank		0.000	05:13
Subject Sample	# 2	0.039	05:14
Breath Volume		3.332	05:14
Air Blank		0.000	05:15
Control Test		0.080	05:15
Air Blank		0.000	05:16
Diagnostics Check		OK	05:16

Data Download Date/Time: 06/16/2017 20:06

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEMEN,AARON,M
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: STEMEN,AARON
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 05/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:28

SUBJECT'S LAST NAME: VORCE

DATE/TIME OF TEST: 06/08/2017 02:05:00

Diagnostics Check		OK	01:56	
Air Blank		0.000	01:56	
Control Test		0.079	01:57	
Air Blank		0.000	01:57	
Subject Sample	# 1		02:00	No Sample Provided
Breath Volume		0.000	02:00	
Air Blank		0.000	02:01	
Air Blank		0.000	02:03	
Subject Sample	# 2		02:03	Subject Test Refused
Breath Volume		0.000	02:03	
Air Blank		0.000	02:04	
Control Test		0.080	02:04	
Air Blank		0.000	02:04	
Diagnostics Check		OK	02:05	

Data Download Date/Time: 06/16/2017 20:06

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEMEN,AARON,M
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: STEMEN,AARON
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 05/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:57

SUBJECT'S LAST NAME: AMONTREE

DATE/TIME OF TEST: 06/09/2017 03:26:49

Diagnostics Check		OK	03:20
Air Blank		0.000	03:20
Control Test		0.080	03:21
Air Blank		0.000	03:21
Subject Sample	# 1	0.140	03:22
Breath Volume		3.644	03:22
Air Blank		0.000	03:22
Air Blank		0.000	03:24
Subject Sample	# 2	0.136	03:25
Breath Volume		3.394	03:25
Air Blank		0.000	03:25
Control Test		0.079	03:26
Air Blank		0.000	03:26
Diagnostics Check		OK	03:26

Data Download Date/Time: 06/16/2017 20:06

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEMEN,AARON,M
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: STEMEN,AARON
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 05/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:21

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 06/10/2017 02:49:33

Diagnostics Check		OK	02:43
Air Blank		0.000	02:43
Control Test		0.080	02:44
Air Blank		0.000	02:44
Subject Sample	# 1	0.119	02:45
Breath Volume		3.699	02:45
Air Blank		0.000	02:45
Air Blank		0.000	02:47
Subject Sample	# 2	0.113	02:48
Breath Volume		3.187	02:48
Air Blank		0.000	02:48
Control Test		0.080	02:48
Air Blank		0.000	02:49
Diagnostics Check		OK	02:49

Data Download Date/Time: 06/16/2017 20:06

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEMEN,AARON,M
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: STEMEN,AARON
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 05/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:16

SUBJECT'S LAST NAME: DANIELS

DATE/TIME OF TEST: 06/10/2017 22:45:48

Diagnostics Check		OK	22:36
Air Blank		0.000	22:36
Control Test		0.079	22:37
Air Blank		0.000	22:37
Subject Sample	# 1	0.314	22:37
Breath Volume		2.734	22:37
Air Blank		0.000	22:38
Air Blank		0.000	22:40
Subject Sample	# 2	0.286	22:41
Breath Volume		2.589	22:41
Air Blank		0.000	22:42
Air Blank		0.000	22:43
Subject Sample	# 3	0.303	22:44
Breath Volume		2.925	22:44
Air Blank		0.000	22:44
Control Test		0.079	22:45
Air Blank		0.000	22:45
Diagnostics Check		OK	22:45

Data Download Date/Time: 06/16/2017 20:07

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOSTICK,JESSE,W
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: KANKAM,DENZEL
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 05/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 12:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 06/12/2017 01:07:03

Diagnostics Check		OK	00:55	
Air Blank		0.000	00:56	
Control Test		0.078	00:56	
Air Blank		0.000	00:56	
Subject Sample	# 1		01:00	No Sample Provided
Breath Volume		0.000	01:00	
Air Blank		0.000	01:00	
Air Blank		0.000	01:02	
Subject Sample	# 2		01:05	No Sample Provided
Breath Volume		0.000	01:05	
Air Blank		0.000	01:06	
Control Test		0.077	01:06	
Air Blank		0.000	01:06	
Diagnostics Check		OK	01:07	

Data Download Date/Time: 06/16/2017 20:07

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SOWARDS,JACOB,T
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: SOWARDS,JACOB
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 05/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:03

SUBJECT'S LAST NAME: TUBBS

DATE/TIME OF TEST: 06/14/2017 01:30:22

Diagnostics Check		OK	01:24
Air Blank		0.000	01:24
Control Test		0.079	01:24
Air Blank		0.000	01:25
Subject Sample	# 1	0.185	01:25
Breath Volume		2.851	01:25
Air Blank		0.000	01:26
Air Blank		0.000	01:28
Subject Sample	# 2	0.192	01:28
Breath Volume		4.062	01:28
Air Blank		0.000	01:29
Control Test		0.078	01:29
Air Blank		0.000	01:30
Diagnostics Check		OK	01:30

Data Download Date/Time: 06/16/2017 20:07

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POSTLE,SHELLEY,M
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: GUYTON,JOHN
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 05/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: WOOD

DATE/TIME OF TEST: 06/15/2017 01:38:15

Diagnostics Check		OK	01:31
Air Blank		0.000	01:32
Control Test		0.080	01:32
Air Blank		0.000	01:32
Subject Sample	# 1	0.187	01:33
Breath Volume		1.996	01:33
Air Blank		0.000	01:34
Air Blank		0.000	01:36
Subject Sample	# 2	0.200	01:36
Breath Volume		2.640	01:36
Air Blank		0.000	01:37
Control Test		0.079	01:37
Air Blank		0.000	01:38
Diagnostics Check		OK	01:38

Data Download Date/Time: 06/16/2017 20:07

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEMEN,AARON,M
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: GUYTON,JOHN
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 05/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:29

SUBJECT'S LAST NAME: PONDS-WHYTUS

DATE/TIME OF TEST: 06/16/2017 03:50:31

Diagnostics Check	OK	03:49	
Air Blank	0.000	03:49	
Control Test	0.077	03:50	Interferent Detect
Air Blank	0.000	03:50	

Data Download Date/Time: 06/16/2017 20:07

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEPHENS,TRAVIS,L
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: STEPHENS,TRAVIS
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 06/16/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:56

SUBJECT'S LAST NAME: COOK

DATE/TIME OF TEST: 06/22/2017 21:29:35

Diagnostics Check		OK	21:16	
Air Blank		0.000	21:17	
Control Test		0.080	21:17	
Air Blank		0.000	21:18	
Subject Sample	# 1	0.184	21:21	Volume Not Met
Breath Volume		1.492	21:21	
Air Blank		0.000	21:22	
Air Blank		0.000	21:23	
Subject Sample	# 2	0.202	21:25	
Breath Volume		1.316	21:25	
Air Blank		0.000	21:25	
Air Blank		0.000	21:27	
Subject Sample	# 3	0.201	21:28	
Breath Volume		1.445	21:28	
Air Blank		0.000	21:28	
Control Test		0.077	21:28	
Air Blank		0.000	21:29	
Diagnostics Check		OK	21:29	

Data Download Date/Time: 07/06/2017 20:17

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POSTLE,SHELLEY,M
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: GUYTON,JOHN
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 06/16/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: BARRERA-COMAS

DATE/TIME OF TEST: 06/24/2017 01:46:11

Diagnostics Check		OK	01:39
Air Blank		0.000	01:40
Control Test		0.080	01:40
Air Blank		0.000	01:41
Subject Sample	# 1	0.177	01:41
Breath Volume		3.125	01:41
Air Blank		0.000	01:42
Air Blank		0.000	01:44
Subject Sample	# 2	0.177	01:44
Breath Volume		3.062	01:44
Air Blank		0.000	01:45
Control Test		0.079	01:45
Air Blank		0.000	01:46
Diagnostics Check		OK	01:46

Data Download Date/Time: 07/06/2017 20:17

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOSTICK,JESSE,W
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: SALAFRIO,JACK
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 06/16/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:12

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 06/25/2017 23:33:50

Diagnostics Check	OK	23:32	
Air Blank	0.000	23:33	
Control Test	0.079	23:33	Interferent Detect
Air Blank	0.000	23:33	

Data Download Date/Time: 07/06/2017 20:17

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOSTICK,JESSE,W
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: SALAFRIO,JACK
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 06/16/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:12

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 06/25/2017 23:36:33

Diagnostics Check	OK	23:35	
Air Blank	0.000	23:35	
Control Test	0.079	23:36	Interferent Detect
Air Blank	0.000	23:36	

Data Download Date/Time: 07/06/2017 20:17

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOSTICK,JESSE,W
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: SALAFRIO,JACK
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 06/16/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:12

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 06/25/2017 23:39:03

Diagnostics Check	OK	23:37	
Air Blank	0.000	23:38	
Control Test	0.079	23:38	Interferent Detect
Air Blank	0.000	23:39	

Data Download Date/Time: 07/06/2017 20:17

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOSTICK,JESSE,W
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: SALAFRIO,JACK
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 06/16/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:12

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 06/25/2017 23:41:32

Diagnostics Check	OK	23:40	
Air Blank	0.000	23:40	
Control Test	0.079	23:41	Interferent Detect
Air Blank	0.000	23:41	

Data Download Date/Time: 07/06/2017 20:17

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001365

INSTRUMENT REGISTERED TO: LEVY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSTON,ROBERT,B
BREATH TEST OPERATOR AGENCY: FFWCC (Lake City)
ARRESTING OFFICER: JOHNSTON,ROBERT
ARRESTING OFFICER AGENCY: FFWCC (Lake City)
LAST AGENCY INSPECTION: 05/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG619605
DRY GAS EXPIRY DATE: 07/14/2018
VIOLATION CODE: BUI
ARREST TIME: 21:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 06/09/2017 22:54:18

Diagnostics Check		OK	22:48
Air Blank		0.000	22:48
Control Test		0.081	22:48
Air Blank		0.000	22:49
Subject Sample	# 1	0.068	22:49
Breath Volume		3.320	22:49
Air Blank		0.000	22:50
Air Blank		0.000	22:52
Subject Sample	# 2	0.063	22:52
Breath Volume		3.101	22:52
Air Blank		0.000	22:53
Control Test		0.081	22:53
Air Blank		0.000	22:54
Diagnostics Check		OK	22:54

Data Download Date/Time: 06/24/2017 14:54

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001365

INSTRUMENT REGISTERED TO: LEVY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANUS,GLEN,A
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: GANUS,GLEN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 05/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG619605
DRY GAS EXPIRY DATE: 07/14/2018
VIOLATION CODE: DUI
ARREST TIME: 01:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:43

SUBJECT'S LAST NAME: MATTHEWS

DATE/TIME OF TEST: 06/12/2017 03:10:28

Diagnostics Check		OK	03:03
Air Blank		0.000	03:03
Control Test		0.081	03:04
Air Blank		0.000	03:04
Subject Sample	# 1	0.148	03:06
Breath Volume		1.546	03:06
Air Blank		0.000	03:06
Air Blank		0.000	03:08
Subject Sample	# 2	0.143	03:08
Breath Volume		1.367	03:08
Air Blank		0.000	03:09
Control Test		0.081	03:09
Air Blank		0.000	03:10
Diagnostics Check		OK	03:10

Data Download Date/Time: 06/24/2017 14:54

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001365

INSTRUMENT REGISTERED TO: LEVY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNYDER,SCOTT,C
BREATH TEST OPERATOR AGENCY: Levy County SO
ARRESTING OFFICER: DAVIS,KEVIN
ARRESTING OFFICER AGENCY: Levy County SO
LAST AGENCY INSPECTION: 05/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG619605
DRY GAS EXPIRY DATE: 07/14/2018
VIOLATION CODE: DUI
ARREST TIME: 01:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:37

SUBJECT'S LAST NAME: RICHARDSON

DATE/TIME OF TEST: 06/18/2017 02:22:51

Diagnostics Check		OK	02:16
Air Blank		0.000	02:16
Control Test		0.081	02:17
Air Blank		0.000	02:17
Subject Sample	# 1	0.115	02:18
Breath Volume		1.175	02:18
Air Blank		0.000	02:19
Air Blank		0.000	02:20
Subject Sample	# 2	0.114	02:21
Breath Volume		2.187	02:21
Air Blank		0.000	02:21
Control Test		0.080	02:22
Air Blank		0.000	02:22
Diagnostics Check		OK	02:22

Data Download Date/Time: 06/24/2017 14:54

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001365

INSTRUMENT REGISTERED TO: LEVY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANDIS,BRYAN,S
BREATH TEST OPERATOR AGENCY: Williston PD
ARRESTING OFFICER: LANDIS,BRYAN
ARRESTING OFFICER AGENCY: Williston PD
LAST AGENCY INSPECTION: 06/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG619605
DRY GAS EXPIRY DATE: 07/14/2018
VIOLATION CODE: DUI
ARREST TIME: 03:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: CURRY

DATE/TIME OF TEST: 06/25/2017 05:04:20

Diagnostics Check		OK	04:58
Air Blank		0.000	04:58
Control Test		0.081	04:59
Air Blank		0.000	04:59
Subject Sample	# 1	0.159	04:59
Breath Volume		1.824	04:59
Air Blank		0.000	05:00
Air Blank		0.000	05:02
Subject Sample	# 2	0.152	05:02
Breath Volume		1.328	05:02
Air Blank		0.000	05:03
Control Test		0.078	05:03
Air Blank		0.000	05:04
Diagnostics Check		OK	05:04

Data Download Date/Time: 07/30/2017 16:37

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PERSAUD,COMPTON,D
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: PERSAUD,COMPTON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 23:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: MCROBERTS

DATE/TIME OF TEST: 06/08/2017 00:22:19

Diagnostics Check		OK	00:15
Air Blank		0.000	00:16
Control Test		0.079	00:16
Air Blank		0.000	00:16
Subject Sample	# 1	0.182	00:17
Breath Volume		1.574	00:17
Air Blank		0.000	00:18
Air Blank		0.000	00:20
Subject Sample	# 2	0.183	00:20
Breath Volume		1.375	00:20
Air Blank		0.000	00:21
Control Test		0.078	00:21
Air Blank		0.000	00:22
Diagnostics Check		OK	00:22

Data Download Date/Time: 06/28/2017 21:40

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDEN,WILLIAM,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HALLOCK,STEPHEN
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 06/14/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: COLLINS

DATE/TIME OF TEST: 06/21/2017 03:27:46

Diagnostics Check		OK	03:21
Air Blank		0.000	03:21
Control Test		0.080	03:22
Air Blank		0.000	03:22
Subject Sample	# 1	0.125	03:23
Breath Volume		2.242	03:23
Air Blank		0.000	03:23
Air Blank		0.000	03:25
Subject Sample	# 2	0.125	03:26
Breath Volume		2.503	03:26
Air Blank		0.000	03:26
Control Test		0.080	03:27
Air Blank		0.000	03:27
Diagnostics Check		OK	03:27

Data Download Date/Time: 06/22/2017 15:00

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDEN,WILLIAM,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: AZEVEDO,FABIO
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 06/14/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: CASTRO

DATE/TIME OF TEST: 06/23/2017 03:27:53

Diagnostics Check		OK	03:21
Air Blank		0.000	03:21
Control Test		0.080	03:21
Air Blank		0.000	03:22
Subject Sample	# 1	0.188	03:23
Breath Volume		1.761	03:23
Air Blank		0.000	03:23
Air Blank		0.000	03:25
Subject Sample	# 2	0.171	03:26
Breath Volume		1.675	03:26
Air Blank		0.000	03:26
Control Test		0.079	03:27
Air Blank		0.000	03:27
Diagnostics Check		OK	03:27

Data Download Date/Time: 06/29/2017 14:52

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO SANTIAGO,ELDALY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DRACH,S
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:09

SUBJECT'S LAST NAME: ALLEN

DATE/TIME OF TEST: 06/01/2017 03:45:17

Diagnostics Check		OK	03:33	
Air Blank		0.000	03:34	
Control Test		0.083	03:34	
Air Blank		0.000	03:35	
Subject Sample	# 1	0.031	03:38	Volume Not Met
Breath Volume		0.160	03:38	
Air Blank		0.000	03:38	
Air Blank		0.000	03:40	
Subject Sample	# 2	0.030	03:43	Volume Not Met
Breath Volume		0.097	03:43	
Air Blank		0.000	03:44	
Control Test		0.082	03:44	
Air Blank		0.000	03:45	
Diagnostics Check		OK	03:45	

Data Download Date/Time: 06/08/2017 14:13

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO SANTIAGO,ELDALY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BAKSH,S
ARRESTING OFFICER AGENCY: Winter Garden PD
LAST AGENCY INSPECTION: 05/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:18

SUBJECT'S LAST NAME: ALVAREZ

DATE/TIME OF TEST: 06/01/2017 22:50:40

Diagnostics Check		OK	22:44
Air Blank		0.000	22:44
Control Test		0.081	22:45
Air Blank		0.000	22:45
Subject Sample	# 1	0.193	22:46
Breath Volume		4.035	22:46
Air Blank		0.000	22:46
Air Blank		0.000	22:48
Subject Sample	# 2	0.191	22:49
Breath Volume		4.281	22:49
Air Blank		0.000	22:49
Control Test		0.082	22:50
Air Blank		0.000	22:50
Diagnostics Check		OK	22:50

Data Download Date/Time: 06/08/2017 14:13

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CANER,OSVALDO,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SIMPSON,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:43

SUBJECT'S LAST NAME: ROSS

DATE/TIME OF TEST: 06/03/2017 02:13:40

Diagnostics Check		OK	02:06
Air Blank		0.000	02:06
Control Test		0.082	02:07
Air Blank		0.000	02:07
Subject Sample	# 1	0.190	02:08
Breath Volume		1.519	02:08
Air Blank		0.000	02:08
Air Blank		0.000	02:10
Subject Sample	# 2	0.190	02:12
Breath Volume		1.941	02:12
Air Blank		0.000	02:12
Control Test		0.081	02:13
Air Blank		0.000	02:13
Diagnostics Check		OK	02:13

Data Download Date/Time: 06/08/2017 14:13

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELBON,NICHOLAS,C
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: FRANCO,CARLOS
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:54

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 06/04/2017 04:22:38

Diagnostics Check		OK	04:16
Air Blank		0.000	04:17
Control Test		0.083	04:17
Air Blank		0.000	04:17
Subject Sample	# 1	0.038	04:18
Breath Volume		1.105	04:18
Air Blank		0.000	04:18
Air Blank		0.000	04:20
Subject Sample	# 2	0.036	04:21
Breath Volume		2.007	04:21
Air Blank		0.000	04:21
Control Test		0.082	04:22
Air Blank		0.000	04:22
Diagnostics Check		OK	04:22

Data Download Date/Time: 06/08/2017 14:13

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELBON,NICHOLAS,C
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: TORRES,NICOLE
ARRESTING OFFICER AGENCY: Windermere PD
LAST AGENCY INSPECTION: 05/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:33

SUBJECT'S LAST NAME: BURGETT

DATE/TIME OF TEST: 06/05/2017 02:04:06

Diagnostics Check		OK	01:57
Air Blank		0.000	01:58
Control Test		0.083	01:58
Air Blank		0.000	01:59
Subject Sample	# 1	0.138	01:59
Breath Volume		1.214	01:59
Air Blank		0.000	02:00
Air Blank		0.000	02:01
Subject Sample	# 2	0.130	02:02
Breath Volume		1.300	02:02
Air Blank		0.000	02:03
Control Test		0.082	02:03
Air Blank		0.000	02:04
Diagnostics Check		OK	02:04

Data Download Date/Time: 06/08/2017 14:13

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CANER,OSVALDO,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: JORGE,JUNIOR
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:53

SUBJECT'S LAST NAME: AVILES-TRIGUEROS

DATE/TIME OF TEST: 06/05/2017 05:22:51

Diagnostics Check		OK	05:16
Air Blank		0.000	05:16
Control Test		0.082	05:17
Air Blank		0.000	05:17
Subject Sample	# 1	0.114	05:18
Breath Volume		3.601	05:18
Air Blank		0.000	05:19
Air Blank		0.000	05:20
Subject Sample	# 2	0.094	05:21
Breath Volume		1.351	05:21
Air Blank		0.000	05:21
Control Test		0.082	05:22
Air Blank		0.000	05:22
Diagnostics Check		OK	05:22

Data Download Date/Time: 06/08/2017 14:13

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELBON,NICHOLAS,C
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ELLER,JESSICA
ARRESTING OFFICER AGENCY: Winter Park PD
LAST AGENCY INSPECTION: 05/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:44

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 06/06/2017 02:13:27

Diagnostics Check		OK	02:07
Air Blank		0.000	02:07
Control Test		0.080	02:07
Air Blank		0.000	02:08
Subject Sample	# 1	0.241	02:08
Breath Volume		2.531	02:08
Air Blank		0.000	02:09
Air Blank		0.000	02:11
Subject Sample	# 2	0.240	02:11
Breath Volume		2.187	02:11
Air Blank		0.000	02:12
Control Test		0.083	02:12
Air Blank		0.000	02:13
Diagnostics Check		OK	02:13

Data Download Date/Time: 06/08/2017 14:13

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELBON,NICHOLAS,C
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ROBLEDO,MIGUEL
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:42

SUBJECT'S LAST NAME: BEAUFORD

DATE/TIME OF TEST: 06/06/2017 04:16:48

Diagnostics Check		OK	04:09
Air Blank		0.000	04:10
Control Test		0.082	04:10
Air Blank		0.000	04:10
Subject Sample	# 1	0.279	04:12
Breath Volume		1.429	04:12
Air Blank		0.000	04:13
Air Blank		0.000	04:14
Subject Sample	# 2	0.294	04:15
Breath Volume		1.125	04:15
Air Blank		0.000	04:15
Control Test		0.082	04:16
Air Blank		0.000	04:16
Diagnostics Check		OK	04:16

Data Download Date/Time: 06/08/2017 14:13

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO SANTIAGO,ELDALY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HALLOCK,S
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 05/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:29

SUBJECT'S LAST NAME: KELLY

DATE/TIME OF TEST: 06/07/2017 03:56:37

Diagnostics Check		OK	03:50
Air Blank		0.000	03:50
Control Test		0.083	03:51
Air Blank		0.000	03:51
Subject Sample	# 1	0.159	03:52
Breath Volume		1.207	03:52
Air Blank		0.000	03:52
Air Blank		0.000	03:54
Subject Sample	# 2	0.151	03:55
Breath Volume		3.105	03:55
Air Blank		0.000	03:55
Control Test		0.082	03:56
Air Blank		0.000	03:56
Diagnostics Check		OK	03:56

Data Download Date/Time: 06/08/2017 14:13

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO SANTIAGO,ELDALY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HAWKINS,D
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:05

SUBJECT'S LAST NAME: DE JESUS

DATE/TIME OF TEST: 06/08/2017 05:32:37

Diagnostics Check		OK	05:26
Air Blank		0.000	05:27
Control Test		0.080	05:27
Air Blank		0.000	05:27
Subject Sample	# 1	0.202	05:28
Breath Volume		3.062	05:28
Air Blank		0.000	05:28
Air Blank		0.000	05:30
Subject Sample	# 2	0.200	05:31
Breath Volume		3.320	05:31
Air Blank		0.000	05:31
Control Test		0.078	05:32
Air Blank		0.000	05:32
Diagnostics Check		OK	05:32

Data Download Date/Time: 06/08/2017 14:13

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO SANTIAGO,ELDALY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DONAKOWSKI,J
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:56

SUBJECT'S LAST NAME: RIGDON

DATE/TIME OF TEST: 06/14/2017 02:27:07

Diagnostics Check		OK	02:20
Air Blank		0.000	02:21
Control Test		0.081	02:21
Air Blank		0.000	02:22
Subject Sample	# 1	0.248	02:22
Breath Volume		1.683	02:22
Air Blank		0.000	02:23
Air Blank		0.000	02:25
Subject Sample	# 2	0.246	02:25
Breath Volume		2.421	02:25
Air Blank		0.000	02:26
Control Test		0.082	02:26
Air Blank		0.000	02:27
Diagnostics Check		OK	02:27

Data Download Date/Time: 06/14/2017 12:53

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO SANTIAGO,ELDALY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HALLOCK,S
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 05/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: DIAS

DATE/TIME OF TEST: 06/14/2017 04:26:39

Diagnostics Check		OK	04:17	
Air Blank		0.000	04:17	
Control Test		0.082	04:18	
Air Blank		0.000	04:18	
Subject Sample	# 1	0.112	04:21	Volume Not Met
Breath Volume		0.695	04:21	
Air Blank		0.000	04:22	
Air Blank		0.000	04:24	
Subject Sample	# 2	0.112	04:24	
Breath Volume		1.488	04:24	
Air Blank			04:25	Purge Fail
Air Blank		0.000	04:26	

Data Download Date/Time: 06/14/2017 12:53

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO SANTIAGO,ELDALY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HALLOCK,S
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 05/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: DIAS

DATE/TIME OF TEST: 06/14/2017 04:29:28

Diagnostics Check	OK	04:28	
Air Blank	0.000	04:28	
Control Test	0.069	04:28	Control Outside Tolerance
Air Blank	0.000	04:29	

Data Download Date/Time: 06/14/2017 12:53

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON,TRAVIS,W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HACKEY,CHRISTIN
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 06/14/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:54

SUBJECT'S LAST NAME: GORE

DATE/TIME OF TEST: 06/14/2017 22:21:10

Diagnostics Check		OK	22:15
Air Blank		0.000	22:15
Control Test		0.082	22:15
Air Blank		0.000	22:16
Subject Sample	# 1	0.049	22:16
Breath Volume		1.281	22:16
Air Blank		0.000	22:17
Air Blank		0.000	22:19
Subject Sample	# 2	0.049	22:19
Breath Volume		2.183	22:19
Air Blank		0.000	22:20
Control Test		0.082	22:20
Air Blank		0.000	22:21
Diagnostics Check		OK	22:21

Data Download Date/Time: 06/22/2017 15:03

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,CALEB,N
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: AZEVADO,FABIO
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 06/14/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:22

SUBJECT'S LAST NAME: NIEVESDEJESUS

DATE/TIME OF TEST: 06/15/2017 03:53:28

Diagnostics Check		OK	03:46
Air Blank		0.000	03:47
Control Test		0.082	03:47
Air Blank		0.000	03:48
Subject Sample	# 1	0.122	03:48
Breath Volume		4.507	03:48
Air Blank		0.000	03:49
Air Blank		0.000	03:51
Subject Sample	# 2	0.112	03:51
Breath Volume		4.460	03:51
Air Blank		0.000	03:52
Control Test		0.080	03:52
Air Blank		0.000	03:53
Diagnostics Check		OK	03:53

Data Download Date/Time: 06/22/2017 15:03

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OKEEFE,ALEXIS,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: WILLIX,DONNA
ARRESTING OFFICER AGENCY: Winter Garden PD
LAST AGENCY INSPECTION: 06/14/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 17:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:30

SUBJECT'S LAST NAME: HENESEY

DATE/TIME OF TEST: 06/15/2017 19:00:15

Diagnostics Check		OK	18:54
Air Blank		0.000	18:54
Control Test		0.082	18:55
Air Blank		0.000	18:55
Subject Sample	# 1	0.253	18:55
Breath Volume		2.015	18:55
Air Blank		0.000	18:56
Air Blank		0.000	18:58
Subject Sample	# 2	0.251	18:58
Breath Volume		1.644	18:58
Air Blank		0.000	18:59
Control Test		0.081	18:59
Air Blank		0.000	19:00
Diagnostics Check		OK	19:00

Data Download Date/Time: 06/22/2017 15:03

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON,TRAVIS,W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MCKENNA,CONOR
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 06/14/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:37

SUBJECT'S LAST NAME: SPROUSE

DATE/TIME OF TEST: 06/16/2017 02:00:52

Diagnostics Check		OK	01:58	
Air Blank		0.000	01:58	
Control Test		0.082	01:58	
Air Blank		0.000	01:59	
Subject Sample	# 1		01:59	Subject Test Refused
Breath Volume		0.000	01:59	
Air Blank		0.000	01:59	
Control Test		0.082	02:00	
Air Blank		0.000	02:00	
Diagnostics Check		OK	02:00	

Data Download Date/Time: 06/22/2017 15:03

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,CALEB,N
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CSIZMADIA,ANDY
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 06/14/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:34

SUBJECT'S LAST NAME: SANTOS

DATE/TIME OF TEST: 06/16/2017 04:04:37

Diagnostics Check		OK	03:57
Air Blank		0.000	03:57
Control Test		0.082	03:58
Air Blank		0.000	03:58
Subject Sample	# 1	0.149	04:00
Breath Volume		2.585	04:00
Air Blank		0.000	04:00
Air Blank		0.000	04:02
Subject Sample	# 2	0.145	04:03
Breath Volume		1.546	04:03
Air Blank		0.000	04:03
Control Test		0.081	04:04
Air Blank		0.000	04:04
Diagnostics Check		OK	04:04

Data Download Date/Time: 06/22/2017 15:03

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON,TRAVIS,W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: LAWRENCE,NICKOLAS
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 06/14/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:18

SUBJECT'S LAST NAME: DEL VECCHIO

DATE/TIME OF TEST: 06/16/2017 04:50:31

Diagnostics Check		OK	04:43
Air Blank		0.000	04:44
Control Test		0.082	04:44
Air Blank		0.000	04:44
Subject Sample	# 1	0.222	04:45
Breath Volume		1.246	04:45
Air Blank		0.000	04:45
Air Blank		0.000	04:47
Subject Sample	# 2	0.227	04:48
Breath Volume		1.363	04:48
Air Blank		0.000	04:49
Control Test		0.081	04:49
Air Blank		0.000	04:50
Diagnostics Check		OK	04:50

Data Download Date/Time: 06/22/2017 15:03

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OKEEFE,ALEXIS,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: WONGSHUE,ANTHONY
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 06/14/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:20

SUBJECT'S LAST NAME: CAMIRE

DATE/TIME OF TEST: 06/16/2017 19:48:41

Diagnostics Check		OK	19:42
Air Blank		0.000	19:42
Control Test		0.081	19:43
Air Blank		0.000	19:43
Subject Sample	# 1	0.277	19:44
Breath Volume		2.867	19:44
Air Blank		0.000	19:44
Air Blank		0.000	19:46
Subject Sample	# 2	0.282	19:47
Breath Volume		2.988	19:47
Air Blank		0.000	19:47
Control Test		0.081	19:48
Air Blank		0.000	19:48
Diagnostics Check		OK	19:48

Data Download Date/Time: 06/22/2017 15:03

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: WEAVER,WILLIAM
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 06/14/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:46

SUBJECT'S LAST NAME: JIN

DATE/TIME OF TEST: 06/17/2017 00:17:01

Diagnostics Check		OK	00:10
Air Blank		0.000	00:10
Control Test		0.081	00:10
Air Blank		0.000	00:11
Subject Sample	# 1	0.158	00:11
Breath Volume		1.531	00:11
Air Blank		0.000	00:12
Air Blank		0.000	00:14
Subject Sample	# 2	0.170	00:15
Breath Volume		2.781	00:15
Air Blank		0.000	00:16
Control Test		0.083	00:16
Air Blank		0.000	00:16
Diagnostics Check		OK	00:17

Data Download Date/Time: 06/22/2017 15:03

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,CALEB,N
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: WONGSHUE,ANTHONY
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 06/14/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:23

SUBJECT'S LAST NAME: ARCE

DATE/TIME OF TEST: 06/17/2017 01:02:25

Diagnostics Check		OK	00:47	
Air Blank		0.000	00:47	
Control Test		0.082	00:47	
Air Blank		0.000	00:48	
Subject Sample	# 1	0.130	00:51	Volume Not Met
Breath Volume		0.636	00:51	
Air Blank		0.000	00:52	
Air Blank		0.000	00:54	
Subject Sample	# 2	0.139	00:55	
Breath Volume		1.343	00:55	
Air Blank		0.000	00:55	
Air Blank		0.000	00:57	
Subject Sample	# 3	0.110	01:00	Volume Not Met
Breath Volume		0.316	01:00	
Air Blank		0.000	01:01	
Control Test		0.081	01:01	
Air Blank		0.000	01:02	
Diagnostics Check		OK	01:02	

Data Download Date/Time: 06/22/2017 15:03

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON,TRAVIS,W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: AZEVEDO,FABIO
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 06/14/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:14

SUBJECT'S LAST NAME: RHODEBACK

DATE/TIME OF TEST: 06/17/2017 02:42:15

Diagnostics Check		OK	02:36
Air Blank		0.000	02:36
Control Test		0.082	02:36
Air Blank		0.000	02:37
Subject Sample	# 1	0.161	02:37
Breath Volume		4.550	02:37
Air Blank		0.000	02:38
Air Blank		0.000	02:40
Subject Sample	# 2	0.162	02:40
Breath Volume		4.339	02:40
Air Blank		0.000	02:41
Control Test		0.081	02:41
Air Blank		0.000	02:42
Diagnostics Check		OK	02:42

Data Download Date/Time: 06/22/2017 15:03

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,CALEB,N
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: YUKNIEWICZ,BRIANA
ARRESTING OFFICER AGENCY: Winter Garden PD
LAST AGENCY INSPECTION: 06/14/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:26

SUBJECT'S LAST NAME: HOWARD

DATE/TIME OF TEST: 06/17/2017 04:00:43

Diagnostics Check		OK	03:51
Air Blank		0.000	03:51
Control Test		0.081	03:51
Air Blank		0.000	03:52
Subject Sample	# 1	0.352	03:53
Breath Volume		1.515	03:53
Air Blank		0.000	03:54
Air Blank		0.000	03:56
Subject Sample	# 2	0.343	03:58
Breath Volume		2.621	03:58
Air Blank		0.000	03:59
Control Test		0.080	04:00
Air Blank		0.000	04:00
Diagnostics Check		OK	04:00

Data Download Date/Time: 06/22/2017 15:03

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OKEEFE,ALEXIS,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BODE,JOSHUA
ARRESTING OFFICER AGENCY: Ocoee PD
LAST AGENCY INSPECTION: 06/14/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: MAKIESE

DATE/TIME OF TEST: 06/18/2017 00:15:38

Diagnostics Check		OK	00:09
Air Blank		0.000	00:09
Control Test		0.082	00:09
Air Blank		0.000	00:10
Subject Sample	# 1	0.121	00:11
Breath Volume		1.453	00:11
Air Blank		0.000	00:11
Air Blank		0.000	00:13
Subject Sample	# 2	0.123	00:14
Breath Volume		1.472	00:14
Air Blank		0.000	00:14
Control Test		0.082	00:15
Air Blank		0.000	00:15
Diagnostics Check		OK	00:15

Data Download Date/Time: 06/22/2017 15:03

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,CALEB,N
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HAWKINS,D
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 06/14/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:32

SUBJECT'S LAST NAME: KELLER

DATE/TIME OF TEST: 06/18/2017 03:11:17

Diagnostics Check		OK	02:59	
Air Blank		0.000	03:00	
Control Test		0.082	03:00	
Air Blank		0.000	03:01	
Subject Sample	# 1		03:04	No Sample Provided
Breath Volume		0.000	03:04	
Air Blank		0.000	03:04	
Air Blank		0.000	03:06	
Subject Sample	# 2		03:09	No Sample Provided
Breath Volume		0.000	03:09	
Air Blank		0.000	03:10	
Control Test		0.082	03:10	
Air Blank		0.000	03:11	
Diagnostics Check		OK	03:11	

Data Download Date/Time: 06/22/2017 15:03

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON,TRAVIS,W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: RODRIGUEZ,STEPHANIE
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 06/14/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: FRAZIER

DATE/TIME OF TEST: 06/18/2017 03:33:44

Diagnostics Check		OK	03:27
Air Blank		0.000	03:27
Control Test		0.081	03:28
Air Blank		0.000	03:28
Subject Sample	# 1	0.144	03:29
Breath Volume		3.476	03:29
Air Blank		0.000	03:29
Air Blank		0.000	03:31
Subject Sample	# 2	0.148	03:32
Breath Volume		3.566	03:32
Air Blank		0.000	03:32
Control Test		0.082	03:33
Air Blank		0.000	03:33
Diagnostics Check		OK	03:33

Data Download Date/Time: 06/22/2017 15:03

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELBON,NICHOLAS,C
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: EDMONDS,STEVEN
ARRESTING OFFICER AGENCY: Winter Garden PD
LAST AGENCY INSPECTION: 06/14/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:48

SUBJECT'S LAST NAME: PUERTOANTUNEZ

DATE/TIME OF TEST: 06/18/2017 04:19:16

Diagnostics Check		OK	04:13
Air Blank		0.000	04:13
Control Test		0.082	04:14
Air Blank		0.000	04:14
Subject Sample	# 1	0.149	04:14
Breath Volume		2.539	04:14
Air Blank		0.000	04:15
Air Blank		0.000	04:17
Subject Sample	# 2	0.149	04:17
Breath Volume		2.476	04:17
Air Blank		0.000	04:18
Control Test		0.082	04:18
Air Blank		0.000	04:19
Diagnostics Check		OK	04:19

Data Download Date/Time: 06/22/2017 15:03

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OKEEFE,ALEXIS,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: WADEMAN,CHRISTOPHER
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 06/14/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:25

SUBJECT'S LAST NAME: GERMAIN

DATE/TIME OF TEST: 06/18/2017 23:03:17

Diagnostics Check		OK	22:51	
Air Blank		0.000	22:52	
Control Test		0.082	22:52	
Air Blank		0.000	22:53	
Subject Sample	# 1		22:56	No Sample Provided
Breath Volume		0.000	22:56	
Air Blank		0.000	22:56	
Air Blank		0.000	22:58	
Subject Sample	# 2		23:01	No Sample Provided
Breath Volume		0.000	23:01	
Air Blank		0.000	23:02	
Control Test		0.083	23:02	
Air Blank		0.000	23:03	
Diagnostics Check		OK	23:03	

Data Download Date/Time: 06/22/2017 15:03

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELBON,NICHOLAS,C
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MATHISEN,LEIGH
ARRESTING OFFICER AGENCY: Winter Garden PD
LAST AGENCY INSPECTION: 06/14/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:01

SUBJECT'S LAST NAME: ADAMS

DATE/TIME OF TEST: 06/20/2017 04:31:48

Diagnostics Check		OK	04:25
Air Blank		0.000	04:25
Control Test		0.082	04:26
Air Blank		0.000	04:26
Subject Sample	# 1	0.091	04:27
Breath Volume		3.617	04:27
Air Blank		0.000	04:27
Air Blank		0.000	04:29
Subject Sample	# 2	0.092	04:30
Breath Volume		3.738	04:30
Air Blank		0.000	04:30
Control Test		0.083	04:31
Air Blank		0.000	04:31
Diagnostics Check		OK	04:31

Data Download Date/Time: 06/22/2017 15:03

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELBON,NICHOLAS,C
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SANTIAGO,WILLIAM
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 06/14/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:00

SUBJECT'S LAST NAME: DORMAN

DATE/TIME OF TEST: 06/20/2017 05:38:01

Diagnostics Check		OK	05:28
Air Blank		0.000	05:29
Control Test		0.082	05:29
Air Blank		0.000	05:30
Subject Sample	# 1	0.192	05:31
Breath Volume		1.714	05:31
Air Blank		0.000	05:32
Air Blank		0.000	05:34
Subject Sample	# 2	0.186	05:36
Breath Volume		1.429	05:36
Air Blank		0.000	05:37
Control Test		0.081	05:37
Air Blank		0.000	05:37
Diagnostics Check		OK	05:38

Data Download Date/Time: 06/22/2017 15:03

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MELVILLE,KELLY,M
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: JORGE,JUNIOR
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 06/14/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 11:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:04

SUBJECT'S LAST NAME: RIOSHOYOS

DATE/TIME OF TEST: 06/21/2017 12:35:48

Diagnostics Check		OK	12:28
Air Blank		0.000	12:28
Control Test		0.083	12:29
Air Blank		0.000	12:29
Subject Sample	# 1	0.000	12:30
Breath Volume		1.843	12:30
Air Blank		0.000	12:30
Air Blank		0.000	12:32
Subject Sample	# 2	0.000	12:34
Breath Volume		1.828	12:34
Air Blank		0.000	12:34
Control Test		0.082	12:35
Air Blank		0.000	12:35
Diagnostics Check		OK	12:35

Data Download Date/Time: 06/22/2017 15:03

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,CALEB,N
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MOONDA,SHUAIB
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 06/14/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:57

SUBJECT'S LAST NAME: QUILICHINIFRADERA

DATE/TIME OF TEST: 06/22/2017 04:28:04

Diagnostics Check		OK	04:21
Air Blank		0.000	04:21
Control Test		0.082	04:21
Air Blank		0.000	04:22
Subject Sample	# 1	0.182	04:23
Breath Volume		1.125	04:23
Air Blank		0.000	04:24
Air Blank		0.000	04:25
Subject Sample	# 2	0.187	04:26
Breath Volume		1.812	04:26
Air Blank		0.000	04:27
Control Test		0.082	04:27
Air Blank		0.000	04:27
Diagnostics Check		OK	04:28

Data Download Date/Time: 06/22/2017 15:03

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HAWKINS,DEB
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 06/14/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 05:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:40

SUBJECT'S LAST NAME: CONKLIN

DATE/TIME OF TEST: 06/22/2017 07:15:26

Diagnostics Check		OK	07:03	
Air Blank		0.000	07:04	
Control Test		0.082	07:04	
Air Blank		0.000	07:05	
Subject Sample	# 1	0.148	07:08	Volume Not Met
Breath Volume		0.585	07:08	
Air Blank		0.000	07:08	
Air Blank		0.000	07:10	
Subject Sample	# 2	0.147	07:13	Volume Not Met
Breath Volume		0.089	07:13	
Air Blank		0.000	07:14	
Control Test		0.082	07:14	
Air Blank		0.000	07:15	
Diagnostics Check		OK	07:15	

Data Download Date/Time: 06/22/2017 15:03

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HAWKINS,DEB
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 06/14/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 05:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:40

SUBJECT'S LAST NAME: CONKLIN

DATE/TIME OF TEST: 06/22/2017 07:25:34

Diagnostics Check		OK	07:17	
Air Blank		0.000	07:17	
Control Test		0.081	07:17	
Air Blank		0.000	07:18	
Subject Sample	# 1	0.138	07:21	Volume Not Met
Breath Volume		0.503	07:21	
Air Blank		0.000	07:22	
Air Blank		0.000	07:23	
Subject Sample	# 2		07:24	Subject Test Refused
Breath Volume		0.000	07:24	
Air Blank		0.000	07:24	
Control Test		0.079	07:24	
Air Blank		0.000	07:25	
Diagnostics Check		OK	07:25	

Data Download Date/Time: 06/22/2017 15:03

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON,TRAVIS,W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MORALES,JOSE
ARRESTING OFFICER AGENCY: Central FL Univ PD
LAST AGENCY INSPECTION: 06/14/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:08

SUBJECT'S LAST NAME: HOLLEY

DATE/TIME OF TEST: 06/23/2017 02:37:54

Diagnostics Check		OK	02:31
Air Blank		0.000	02:31
Control Test		0.083	02:32
Air Blank		0.000	02:32
Subject Sample	# 1	0.115	02:33
Breath Volume		3.105	02:33
Air Blank		0.000	02:34
Air Blank		0.000	02:35
Subject Sample	# 2	0.108	02:36
Breath Volume		1.664	02:36
Air Blank		0.000	02:37
Control Test		0.081	02:37
Air Blank		0.000	02:37
Diagnostics Check		OK	02:37

Data Download Date/Time: 06/29/2017 14:50

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OKEEFE,ALEXIS,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HAWKINS,D
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 06/14/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:05

SUBJECT'S LAST NAME: BROOKS

DATE/TIME OF TEST: 06/23/2017 23:36:33

Diagnostics Check		OK	23:29
Air Blank		0.000	23:29
Control Test		0.083	23:29
Air Blank		0.000	23:30
Subject Sample	# 1	0.204	23:31
Breath Volume		2.949	23:31
Air Blank		0.000	23:31
Air Blank		0.000	23:33
Subject Sample	# 2	0.207	23:34
Breath Volume		1.718	23:34
Air Blank		0.000	23:35
Control Test		0.080	23:35
Air Blank		0.000	23:36
Diagnostics Check		OK	23:36

Data Download Date/Time: 06/29/2017 14:50

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,CALEB,N
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: WILSON,RUDOLPH
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 06/14/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:48

SUBJECT'S LAST NAME: MELOY

DATE/TIME OF TEST: 06/24/2017 01:19:11

Diagnostics Check		OK	01:11
Air Blank		0.000	01:12
Control Test		0.082	01:12
Air Blank		0.000	01:12
Subject Sample	# 1	0.211	01:13
Breath Volume		2.652	01:13
Air Blank		0.000	01:14
Air Blank		0.000	01:16
Subject Sample	# 2	0.213	01:17
Breath Volume		3.093	01:17
Air Blank		0.000	01:18
Control Test		0.081	01:18
Air Blank		0.000	01:19
Diagnostics Check		OK	01:19

Data Download Date/Time: 06/29/2017 14:50

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON,TRAVIS,W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: KERN,ETHAN
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 06/14/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:44

SUBJECT'S LAST NAME: BERRIOS RODRIGUEZ

DATE/TIME OF TEST: 06/24/2017 03:10:45

Diagnostics Check		OK	03:04
Air Blank		0.000	03:05
Control Test		0.082	03:05
Air Blank		0.000	03:05
Subject Sample	# 1	0.177	03:06
Breath Volume		2.906	03:06
Air Blank		0.000	03:06
Air Blank		0.000	03:08
Subject Sample	# 2	0.174	03:09
Breath Volume		2.531	03:09
Air Blank		0.000	03:09
Control Test		0.081	03:10
Air Blank		0.000	03:10
Diagnostics Check		OK	03:10

Data Download Date/Time: 06/29/2017 14:50

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OKEEFE,ALEXIS,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: AZEVEDO,F
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 06/14/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: RHOADES

DATE/TIME OF TEST: 06/25/2017 03:30:14

Diagnostics Check		OK	03:23
Air Blank		0.000	03:24
Control Test		0.083	03:24
Air Blank		0.000	03:25
Subject Sample	# 1	0.212	03:25
Breath Volume		2.542	03:25
Air Blank		0.000	03:26
Air Blank		0.000	03:28
Subject Sample	# 2	0.215	03:28
Breath Volume		3.097	03:28
Air Blank		0.000	03:29
Control Test		0.082	03:29
Air Blank		0.000	03:30
Diagnostics Check		OK	03:30

Data Download Date/Time: 06/29/2017 14:50

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,CALEB,N
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: WINBURN,TAYLOR
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 06/14/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:58

SUBJECT'S LAST NAME: LASANTA MARRERO

DATE/TIME OF TEST: 06/25/2017 04:28:03

Diagnostics Check		OK	04:20
Air Blank		0.000	04:21
Control Test		0.082	04:21
Air Blank		0.000	04:22
Subject Sample	# 1	0.204	04:23
Breath Volume		1.621	04:23
Air Blank		0.000	04:24
Air Blank		0.000	04:25
Subject Sample	# 2	0.190	04:26
Breath Volume		1.847	04:26
Air Blank		0.000	04:27
Control Test		0.081	04:27
Air Blank		0.000	04:27
Diagnostics Check		OK	04:28

Data Download Date/Time: 06/29/2017 14:50

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELBON,NICHOLAS,C
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: PUPO,ANDREW
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 06/14/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:56

SUBJECT'S LAST NAME: MIRAMONTES MENDEZ

DATE/TIME OF TEST: 06/25/2017 05:27:57

Diagnostics Check		OK	05:20
Air Blank		0.000	05:21
Control Test		0.082	05:21
Air Blank		0.000	05:21
Subject Sample	# 1	0.103	05:23
Breath Volume		1.109	05:23
Air Blank		0.000	05:23
Air Blank		0.000	05:25
Subject Sample	# 2	0.096	05:26
Breath Volume		2.101	05:26
Air Blank		0.000	05:27
Control Test		0.082	05:27
Air Blank		0.000	05:27
Diagnostics Check		OK	05:27

Data Download Date/Time: 06/29/2017 14:50

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OKEEFE,ALEXIS,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SIMPSON,J
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 06/14/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 17:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:35

SUBJECT'S LAST NAME: SHIPLEY

DATE/TIME OF TEST: 06/25/2017 19:02:48

Diagnostics Check		OK	18:56
Air Blank		0.000	18:57
Control Test		0.083	18:57
Air Blank		0.000	18:57
Subject Sample	# 1	0.212	18:58
Breath Volume		2.800	18:58
Air Blank		0.000	18:59
Air Blank		0.000	19:00
Subject Sample	# 2	0.213	19:01
Breath Volume		3.058	19:01
Air Blank		0.000	19:01
Control Test		0.081	19:02
Air Blank		0.000	19:02
Diagnostics Check		OK	19:02

Data Download Date/Time: 06/29/2017 14:50

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CANER,OSVALDO,3
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MERKS,ANDREW
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 06/14/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:53

SUBJECT'S LAST NAME: BLANDIN

DATE/TIME OF TEST: 06/25/2017 21:21:19

Diagnostics Check		OK	21:14
Air Blank		0.000	21:15
Control Test		0.083	21:15
Air Blank		0.000	21:15
Subject Sample	# 1	0.202	21:16
Breath Volume		1.449	21:16
Air Blank		0.000	21:17
Air Blank		0.000	21:19
Subject Sample	# 2	0.208	21:19
Breath Volume		2.375	21:19
Air Blank		0.000	21:20
Control Test		0.082	21:20
Air Blank		0.000	21:21
Diagnostics Check		OK	21:21

Data Download Date/Time: 06/29/2017 14:50

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELBON,NICHOLAS,C
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HAINES,BRANT
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 06/14/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: CHRISTENSEN

DATE/TIME OF TEST: 06/26/2017 01:56:33

Diagnostics Check		OK	01:46
Air Blank		0.000	01:47
Control Test		0.082	01:47
Air Blank		0.000	01:48
Subject Sample	# 1	0.042	01:49
Breath Volume		1.195	01:49
Air Blank		0.000	01:50
Air Blank		0.000	01:52
Subject Sample	# 2	0.046	01:55
Breath Volume		2.148	01:55
Air Blank		0.000	01:55
Control Test		0.082	01:55
Air Blank		0.000	01:56
Diagnostics Check		OK	01:56

Data Download Date/Time: 06/29/2017 14:50

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OKEEFE,ALEXIS,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: GRIFFITH,C
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 06/14/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: COCONEHERNANDEZ

DATE/TIME OF TEST: 06/26/2017 03:07:49

Diagnostics Check		OK	03:00
Air Blank		0.000	03:01
Control Test		0.082	03:01
Air Blank		0.000	03:02
Subject Sample	# 1	0.227	03:02
Breath Volume		1.285	03:02
Air Blank		0.000	03:03
Air Blank		0.000	03:05
Subject Sample	# 2	0.228	03:06
Breath Volume		1.988	03:06
Air Blank		0.000	03:06
Control Test		0.082	03:07
Air Blank		0.000	03:07
Diagnostics Check		OK	03:07

Data Download Date/Time: 06/29/2017 14:50

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON,TRAVIS,W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: REYNOLDS,CONNARY
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 06/14/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:31

SUBJECT'S LAST NAME: LONG

DATE/TIME OF TEST: 06/28/2017 21:00:53

Diagnostics Check		OK	20:55
Air Blank		0.000	20:55
Control Test		0.082	20:55
Air Blank		0.000	20:56
Subject Sample	# 1	0.156	20:56
Breath Volume		1.578	20:56
Air Blank		0.000	20:57
Air Blank		0.000	20:59
Subject Sample	# 2	0.163	20:59
Breath Volume		2.328	20:59
Air Blank		0.000	20:59
Control Test		0.082	21:00
Air Blank		0.000	21:00
Diagnostics Check		OK	21:00

Data Download Date/Time: 06/29/2017 14:50

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,CALEB,N
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: AZEVEDO,FABIO
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 06/14/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: STONE

DATE/TIME OF TEST: 06/30/2017 03:19:17

Diagnostics Check		OK	03:12
Air Blank		0.000	03:13
Control Test		0.082	03:13
Air Blank		0.000	03:14
Subject Sample	# 1	0.144	03:14
Breath Volume		2.855	03:14
Air Blank		0.000	03:15
Air Blank		0.000	03:17
Subject Sample	# 2	0.142	03:17
Breath Volume		3.128	03:17
Air Blank		0.000	03:18
Control Test		0.082	03:18
Air Blank		0.000	03:19
Diagnostics Check		OK	03:19

Data Download Date/Time: 07/07/2017 12:57

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON,TRAVIS,W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DRACH,STEFAN
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 06/14/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:00

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 06/30/2017 05:32:00

Diagnostics Check		OK	05:23
Air Blank		0.000	05:23
Control Test		0.082	05:24
Air Blank		0.000	05:24
Subject Sample	# 1	0.172	05:24
Breath Volume		2.054	05:24
Air Blank		0.000	05:25
Air Blank		0.000	05:27
Subject Sample	# 2	0.165	05:30
Breath Volume		1.222	05:30
Air Blank		0.000	05:31
Control Test		0.083	05:31
Air Blank		0.000	05:31
Diagnostics Check		OK	05:32

Data Download Date/Time: 07/07/2017 12:57

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MEADE,CHRISTOPHER
ARRESTING OFFICER AGENCY: Edgewood PD
LAST AGENCY INSPECTION: 06/14/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 05:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:07

SUBJECT'S LAST NAME: COBIANJIMENEZ

DATE/TIME OF TEST: 06/30/2017 07:37:19

Diagnostics Check		OK	07:30
Air Blank		0.000	07:31
Control Test		0.082	07:31
Air Blank		0.000	07:32
Subject Sample	# 1	0.038	07:32
Breath Volume		3.253	07:32
Air Blank		0.000	07:33
Air Blank		0.000	07:35
Subject Sample	# 2	0.038	07:35
Breath Volume		3.804	07:35
Air Blank		0.000	07:36
Control Test		0.082	07:36
Air Blank		0.000	07:37
Diagnostics Check		OK	07:37

Data Download Date/Time: 07/07/2017 12:58

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001442

INSTRUMENT REGISTERED TO: BISCAYNE NAT PARK
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HELMERS,DANIEL,W
BREATH TEST OPERATOR AGENCY: Biscayne Nat. Park
ARRESTING OFFICER: FRAZIER,AMTHONY
ARRESTING OFFICER AGENCY: Biscayne Nat. Park
LAST AGENCY INSPECTION: 05/23/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: BUI
ARREST TIME: 16:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:22

SUBJECT'S LAST NAME: GONZALEZ-ALEAGA

DATE/TIME OF TEST: 06/10/2017 16:56:32

Diagnostics Check		OK	16:47
Air Blank		0.000	16:47
Control Test		0.078	16:48
Air Blank		0.000	16:48
Subject Sample	# 1	0.105	16:51
Breath Volume		2.464	16:51
Air Blank		0.000	16:51
Air Blank		0.000	16:53
Subject Sample	# 2	0.108	16:54
Breath Volume		1.726	16:54
Air Blank		0.000	16:55
Control Test		0.077	16:55
Air Blank		0.000	16:56
Diagnostics Check		OK	16:56

Data Download Date/Time: 06/22/2017 15:06

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001442

INSTRUMENT REGISTERED TO: BISCAYNE NAT PARK
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANTHONY,C
BREATH TEST OPERATOR AGENCY: Biscayne Nat. Park
ARRESTING OFFICER: FRAZIER,ANTHONY
ARRESTING OFFICER AGENCY: Biscayne Nat. Park
LAST AGENCY INSPECTION: 06/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 18:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:31

SUBJECT'S LAST NAME: HERNANDEZ-HERNANDEZ

DATE/TIME OF TEST: 06/25/2017 19:09:20

Diagnostics Check		OK	19:01
Air Blank		0.000	19:01
Control Test		0.078	19:02
Air Blank		0.000	19:02
Subject Sample	# 1	0.147	19:04
Breath Volume		1.812	19:04
Air Blank		0.000	19:05
Air Blank		0.000	19:06
Subject Sample	# 2	0.159	19:07
Breath Volume		1.488	19:07
Air Blank		0.000	19:08
Control Test		0.076	19:08
Air Blank		0.000	19:09
Diagnostics Check		OK	19:09

Data Download Date/Time: 07/28/2017 13:08

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CROOKS,E,T
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: FENNICK,T
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 05/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 14:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:31

SUBJECT'S LAST NAME: MINTON

DATE/TIME OF TEST: 06/04/2017 15:04:05

Diagnostics Check		OK	14:57
Air Blank		0.000	14:58
Control Test		0.080	14:58
Air Blank		0.000	14:58
Subject Sample	# 1	0.315	14:59
Breath Volume		2.675	14:59
Air Blank		0.000	15:00
Air Blank		0.000	15:01
Subject Sample	# 2	0.311	15:02
Breath Volume		2.871	15:02
Air Blank		0.000	15:03
Control Test		0.080	15:03
Air Blank		0.000	15:04
Diagnostics Check		OK	15:04

Data Download Date/Time: 06/18/2017 12:29

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINSON,C,A
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: FIELDS,K
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 05/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: CHILDRESS

DATE/TIME OF TEST: 06/06/2017 23:26:34

Diagnostics Check		OK	23:23	
Air Blank		0.000	23:23	
Control Test		0.081	23:24	
Air Blank		0.000	23:24	
Subject Sample	# 1	0.204	23:25	
Breath Volume		1.601	23:25	
Air Blank			23:25	Improper Sample
Air Blank		0.000	23:25	
Control Test		0.080	23:25	
Air Blank		0.000	23:26	
Diagnostics Check		OK	23:26	

Data Download Date/Time: 06/18/2017 12:29

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINSON,C,A
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: FIELDS,K
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 05/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: CHILDRESS

DATE/TIME OF TEST: 06/06/2017 23:35:23

Diagnostics Check		OK	23:29
Air Blank		0.000	23:29
Control Test		0.080	23:29
Air Blank		0.000	23:30
Subject Sample	# 1	0.207	23:30
Breath Volume		3.171	23:30
Air Blank		0.000	23:31
Air Blank		0.000	23:33
Subject Sample	# 2	0.195	23:33
Breath Volume		1.863	23:33
Air Blank		0.000	23:34
Control Test		0.080	23:34
Air Blank		0.000	23:35
Diagnostics Check		OK	23:35

Data Download Date/Time: 06/18/2017 12:29

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANTORO,CAMERON,J
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: SENAK,K
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 05/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: BURRUS

DATE/TIME OF TEST: 06/07/2017 04:28:59

Diagnostics Check		OK	04:22
Air Blank		0.000	04:23
Control Test		0.081	04:23
Air Blank		0.000	04:23
Subject Sample	# 1	0.152	04:24
Breath Volume		3.722	04:24
Air Blank		0.000	04:24
Air Blank		0.000	04:26
Subject Sample	# 2	0.148	04:27
Breath Volume		3.539	04:27
Air Blank		0.000	04:28
Control Test		0.079	04:28
Air Blank		0.000	04:28
Diagnostics Check		OK	04:28

Data Download Date/Time: 06/18/2017 12:29

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANTORO,CAMERON,J
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: CORDERIO,E
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 05/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 18:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:40

SUBJECT'S LAST NAME: GINDLER

DATE/TIME OF TEST: 06/10/2017 20:20:24

Diagnostics Check		OK	20:14
Air Blank		0.000	20:14
Control Test		0.080	20:15
Air Blank		0.000	20:15
Subject Sample	# 1	0.207	20:16
Breath Volume		1.269	20:16
Air Blank		0.000	20:16
Air Blank		0.000	20:18
Subject Sample	# 2	0.202	20:18
Breath Volume		1.585	20:18
Air Blank		0.000	20:19
Control Test		0.079	20:19
Air Blank		0.000	20:20
Diagnostics Check		OK	20:20

Data Download Date/Time: 06/18/2017 12:29

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001646

INSTRUMENT REGISTERED TO: TYNDALL AFB
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSON,CHRISTOPHER,R
BREATH TEST OPERATOR AGENCY: Tyndall AFB PD
ARRESTING OFFICER: JOHNSON,CHRISTOPHER
ARRESTING OFFICER AGENCY: Tyndall AFB PD
LAST AGENCY INSPECTION: 06/03/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:37

SUBJECT'S LAST NAME: ZAPPITELLA

DATE/TIME OF TEST: 06/10/2017 03:03:54

Diagnostics Check		OK	02:57
Air Blank		0.000	02:58
Control Test		0.080	02:58
Air Blank		0.000	02:59
Subject Sample	# 1	0.000	02:59
Breath Volume		2.593	02:59
Air Blank		0.000	03:00
Air Blank		0.000	03:02
Subject Sample	# 2	0.000	03:02
Breath Volume		2.910	03:02
Air Blank		0.000	03:02
Control Test		0.079	03:03
Air Blank		0.000	03:03
Diagnostics Check		OK	03:03

Data Download Date/Time: 07/24/2017 17:20

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001646

INSTRUMENT REGISTERED TO: TYNDALL AFB
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBER,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Tyndall AFB PD
ARRESTING OFFICER: WEBER,JEFFREY
ARRESTING OFFICER AGENCY: Tyndall AFB PD
LAST AGENCY INSPECTION: 06/03/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:07

SUBJECT'S LAST NAME: WILLIAMSON

DATE/TIME OF TEST: 06/10/2017 22:35:24

Diagnostics Check		OK	22:28
Air Blank		0.000	22:29
Control Test		0.079	22:29
Air Blank		0.000	22:30
Subject Sample	# 1	0.192	22:30
Breath Volume		2.843	22:30
Air Blank		0.000	22:31
Air Blank		0.000	22:33
Subject Sample	# 2	0.195	22:33
Breath Volume		2.421	22:33
Air Blank		0.000	22:34
Control Test		0.077	22:34
Air Blank		0.000	22:35
Diagnostics Check		OK	22:35

Data Download Date/Time: 07/24/2017 17:20

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001646

INSTRUMENT REGISTERED TO: TYNDALL AFB
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSON,CHRISTOPHER,R
BREATH TEST OPERATOR AGENCY: Tyndall AFB PD
ARRESTING OFFICER: WEBER,JEFFREY
ARRESTING OFFICER AGENCY: Tyndall AFB PD
LAST AGENCY INSPECTION: 06/03/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: ADMIN
ARREST TIME: 00:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:22

SUBJECT'S LAST NAME: SULLIVAN

DATE/TIME OF TEST: 06/15/2017 00:52:30

Diagnostics Check		OK	00:46
Air Blank		0.000	00:46
Control Test		0.079	00:47
Air Blank		0.000	00:47
Subject Sample	# 1	0.000	00:48
Breath Volume		1.375	00:48
Air Blank		0.000	00:48
Air Blank		0.000	00:50
Subject Sample	# 2	0.000	00:51
Breath Volume		1.261	00:51
Air Blank		0.000	00:51
Control Test		0.077	00:51
Air Blank		0.000	00:52
Diagnostics Check		OK	00:52

Data Download Date/Time: 07/24/2017 17:20

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001646

INSTRUMENT REGISTERED TO: TYNDALL AFB
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSON,CHRISTOPHER,R
BREATH TEST OPERATOR AGENCY: Tyndall AFB PD
ARRESTING OFFICER: JOHNSON,CHRISTOPHER
ARRESTING OFFICER AGENCY: Tyndall AFB PD
LAST AGENCY INSPECTION: 06/03/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: GLADDING

DATE/TIME OF TEST: 06/16/2017 02:27:47

Diagnostics Check		OK	02:20
Air Blank		0.000	02:21
Control Test		0.079	02:21
Air Blank		0.000	02:22
Subject Sample	# 1	0.107	02:22
Breath Volume		2.363	02:22
Air Blank		0.000	02:23
Air Blank		0.000	02:25
Subject Sample	# 2	0.102	02:26
Breath Volume		3.527	02:26
Air Blank		0.000	02:26
Control Test		0.077	02:27
Air Blank		0.000	02:27
Diagnostics Check		OK	02:27

Data Download Date/Time: 07/24/2017 17:20

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001646

INSTRUMENT REGISTERED TO: TYNDALL AFB
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BATES,JOHNATHON,T
BREATH TEST OPERATOR AGENCY: Tyndall AFB PD
ARRESTING OFFICER: ROBERTSON,COREY
ARRESTING OFFICER AGENCY: Tyndall AFB PD
LAST AGENCY INSPECTION: 06/03/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: ADMIN
ARREST TIME: 22:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:38

SUBJECT'S LAST NAME: ROBERTS

DATE/TIME OF TEST: 06/19/2017 23:10:24

Diagnostics Check		OK	23:03
Air Blank		0.000	23:04
Control Test		0.079	23:04
Air Blank		0.000	23:05
Subject Sample	# 1	0.036	23:05
Breath Volume		1.648	23:05
Air Blank		0.000	23:06
Air Blank		0.000	23:08
Subject Sample	# 2	0.034	23:08
Breath Volume		1.351	23:08
Air Blank		0.000	23:09
Control Test		0.077	23:09
Air Blank		0.000	23:10
Diagnostics Check		OK	23:10

Data Download Date/Time: 07/24/2017 17:20

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001646

INSTRUMENT REGISTERED TO: TYNDALL AFB
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBER,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Tyndall AFB PD
ARRESTING OFFICER: WEBER,JEFFREY
ARRESTING OFFICER AGENCY: Tyndall AFB PD
LAST AGENCY INSPECTION: 06/03/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:48

SUBJECT'S LAST NAME: HUNT

DATE/TIME OF TEST: 06/25/2017 05:16:06

Diagnostics Check		OK	05:09
Air Blank		0.000	05:09
Control Test		0.079	05:10
Air Blank		0.000	05:10
Subject Sample	# 1	0.123	05:11
Breath Volume		2.230	05:11
Air Blank		0.000	05:12
Air Blank		0.000	05:13
Subject Sample	# 2	0.107	05:14
Breath Volume		1.347	05:14
Air Blank		0.000	05:15
Control Test		0.076	05:15
Air Blank		0.000	05:16
Diagnostics Check		OK	05:16

Data Download Date/Time: 07/24/2017 17:20

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001646

INSTRUMENT REGISTERED TO: TYNDALL AFB
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARZA,IVAN,A
BREATH TEST OPERATOR AGENCY: Tyndall AFB PD
ARRESTING OFFICER: GARZA,IVAN
ARRESTING OFFICER AGENCY: Tyndall AFB PD
LAST AGENCY INSPECTION: 06/03/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 15:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:31

SUBJECT'S LAST NAME: STALNAKER

DATE/TIME OF TEST: 06/25/2017 15:58:52

Diagnostics Check		OK	15:51
Air Blank		0.000	15:52
Control Test		0.079	15:52
Air Blank		0.000	15:53
Subject Sample	# 1	0.042	15:54
Breath Volume		1.941	15:54
Air Blank		0.000	15:54
Air Blank		0.000	15:56
Subject Sample	# 2	0.042	15:57
Breath Volume		1.683	15:57
Air Blank		0.000	15:57
Control Test		0.077	15:58
Air Blank		0.000	15:58
Diagnostics Check		OK	15:58

Data Download Date/Time: 07/24/2017 17:20

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001646

INSTRUMENT REGISTERED TO: TYNDALL AFB
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSON,CHRISTOPHER,R
BREATH TEST OPERATOR AGENCY: Tyndall AFB PD
ARRESTING OFFICER: JOHNSON,CHRISTOPHER
ARRESTING OFFICER AGENCY: Tyndall AFB PD
LAST AGENCY INSPECTION: 06/03/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:06

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 06/26/2017 01:32:28

Diagnostics Check		OK	01:26
Air Blank		0.000	01:26
Control Test		0.079	01:27
Air Blank		0.000	01:27
Subject Sample	# 1	0.112	01:27
Breath Volume		2.328	01:27
Air Blank		0.000	01:28
Air Blank		0.000	01:30
Subject Sample	# 2	0.109	01:30
Breath Volume		2.207	01:30
Air Blank		0.000	01:31
Control Test		0.077	01:31
Air Blank		0.000	01:32
Diagnostics Check		OK	01:32

Data Download Date/Time: 07/24/2017 17:20

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001646

INSTRUMENT REGISTERED TO: TYNDALL AFB
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSON,CHRISTOPHER,R
BREATH TEST OPERATOR AGENCY: Tyndall AFB PD
ARRESTING OFFICER: JOHNSON,CHRISTOPHER
ARRESTING OFFICER AGENCY: Tyndall AFB PD
LAST AGENCY INSPECTION: 06/03/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:19

SUBJECT'S LAST NAME: CRONIN

DATE/TIME OF TEST: 06/29/2017 23:46:54

Diagnostics Check		OK	23:39	
Air Blank		0.000	23:39	
Control Test		0.079	23:40	
Air Blank		0.000	23:40	
Subject Sample	# 1	0.089	23:41	
Breath Volume		1.726	23:41	
Air Blank		0.000	23:42	
Air Blank		0.000	23:43	
Subject Sample	# 2	0.083	23:45	
Breath Volume		1.449	23:45	
Air Blank			23:46	RFI Detect
Air Blank		0.000	23:46	

Data Download Date/Time: 07/24/2017 17:20

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TUCKER,HAILEY,E
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: SCARLATO,FRANK
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 05/17/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: LEON

DATE/TIME OF TEST: 06/01/2017 04:27:47

Diagnostics Check		OK	04:20
Air Blank		0.000	04:21
Control Test		0.080	04:21
Air Blank		0.000	04:21
Subject Sample	# 1	0.143	04:23
Breath Volume		2.070	04:23
Air Blank		0.000	04:23
Air Blank		0.000	04:25
Subject Sample	# 2	0.141	04:26
Breath Volume		1.363	04:26
Air Blank		0.000	04:26
Control Test		0.079	04:27
Air Blank		0.000	04:27
Diagnostics Check		OK	04:27

Data Download Date/Time: 06/02/2017 09:50

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARKLEY,JEFFREY,G
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: BARKLEY,JEFFREY
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23016080A3
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:46

SUBJECT'S LAST NAME: KOHL

DATE/TIME OF TEST: 06/03/2017 02:17:18

Diagnostics Check		OK	02:10
Air Blank		0.000	02:10
Control Test		0.079	02:11
Air Blank		0.000	02:11
Subject Sample	# 1	0.136	02:12
Breath Volume		2.136	02:12
Air Blank		0.000	02:13
Air Blank		0.000	02:15
Subject Sample	# 2	0.131	02:15
Breath Volume		1.652	02:15
Air Blank		0.000	02:16
Control Test		0.078	02:16
Air Blank		0.000	02:17
Diagnostics Check		OK	02:17

Data Download Date/Time: 09/26/2017 13:25

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FESSEL,DANIEL,C
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: JONES,NICHOLAS
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: MONTERO

DATE/TIME OF TEST: 06/04/2017 01:06:41

Diagnostics Check		OK	01:00
Air Blank		0.000	01:01
Control Test		0.079	01:01
Air Blank		0.000	01:01
Subject Sample	# 1	0.107	01:02
Breath Volume		2.199	01:02
Air Blank		0.000	01:02
Air Blank		0.000	01:04
Subject Sample	# 2	0.095	01:05
Breath Volume		1.480	01:05
Air Blank		0.000	01:05
Control Test		0.078	01:06
Air Blank		0.000	01:06
Diagnostics Check		OK	01:06

Data Download Date/Time: 09/26/2017 13:25

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEMOS,MATHEW,T
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: RICH,ERICA
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 16:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:43

SUBJECT'S LAST NAME: GALLUZZO

DATE/TIME OF TEST: 06/06/2017 17:10:59

Diagnostics Check		OK	17:04
Air Blank		0.000	17:05
Control Test		0.079	17:05
Air Blank		0.000	17:05
Subject Sample	# 1	0.139	17:06
Breath Volume		3.105	17:06
Air Blank		0.000	17:07
Air Blank		0.000	17:08
Subject Sample	# 2	0.136	17:09
Breath Volume		2.070	17:09
Air Blank		0.000	17:10
Control Test		0.079	17:10
Air Blank		0.000	17:10
Diagnostics Check		OK	17:10

Data Download Date/Time: 09/26/2017 13:25

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001649

INSTRUMENT REGISTERED TO: KENNETH CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FUGATE,JOHN,
BREATH TEST OPERATOR AGENCY: Kenneth City PD
ARRESTING OFFICER: BELTRAN-HERNANDEZ,JOSUE
ARRESTING OFFICER AGENCY: Kenneth City PD
LAST AGENCY INSPECTION: 05/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A3
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 06/10/2017 02:05:20

Diagnostics Check		OK	01:53	
Air Blank		0.000	01:53	
Control Test		0.078	01:53	
Air Blank		0.000	01:54	
Subject Sample	# 1	0.021	01:55	
Breath Volume		2.089	01:55	
Air Blank		0.000	01:55	
Air Blank		0.000	01:57	
Subject Sample	# 2	0.025	02:00	Volume Not Met
Breath Volume		0.136	02:00	
Air Blank		0.000	02:01	
Air Blank		0.000	02:03	
Subject Sample	# 3	0.017	02:03	
Breath Volume		2.324	02:03	
Air Blank		0.000	02:04	
Control Test		0.077	02:04	
Air Blank		0.000	02:05	
Diagnostics Check		OK	02:05	

Data Download Date/Time: 06/22/2017 21:59

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001649

INSTRUMENT REGISTERED TO: KENNETH CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRUNER,SEAN,R
BREATH TEST OPERATOR AGENCY: Kenneth City PD
ARRESTING OFFICER: BRUNER,SEAN
ARRESTING OFFICER AGENCY: Kenneth City PD
LAST AGENCY INSPECTION: 05/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A3
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:49

SUBJECT'S LAST NAME: RABELO

DATE/TIME OF TEST: 06/11/2017 23:23:42

Diagnostics Check		OK	23:14	
Air Blank		0.000	23:15	
Control Test		0.079	23:15	
Air Blank		0.000	23:16	
Subject Sample	# 1		23:19	No Sample Provided
Breath Volume		0.000	23:19	
Air Blank		0.000	23:19	
Air Blank		0.000	23:21	
Subject Sample	# 2		23:22	Subject Test Refused
Breath Volume		0.000	23:22	
Air Blank		0.000	23:22	
Control Test		0.080	23:23	
Air Blank		0.000	23:23	
Diagnostics Check		OK	23:23	

Data Download Date/Time: 06/22/2017 21:59

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001649

INSTRUMENT REGISTERED TO: KENNETH CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRUNER,SEAN,R
BREATH TEST OPERATOR AGENCY: Kenneth City PD
ARRESTING OFFICER: BRUNER,SEAN
ARRESTING OFFICER AGENCY: Kenneth City PD
LAST AGENCY INSPECTION: 05/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A3
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:25

SUBJECT'S LAST NAME: SMAILAGIC

DATE/TIME OF TEST: 06/14/2017 21:05:12

Diagnostics Check		OK	20:58
Air Blank		0.000	20:58
Control Test		0.078	20:58
Air Blank		0.000	20:59
Subject Sample	# 1	0.143	21:00
Breath Volume		2.253	21:00
Air Blank		0.000	21:01
Air Blank		0.000	21:02
Subject Sample	# 2	0.141	21:03
Breath Volume		2.902	21:03
Air Blank		0.000	21:04
Control Test		0.079	21:04
Air Blank		0.000	21:05
Diagnostics Check		OK	21:05

Data Download Date/Time: 06/22/2017 21:59

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001649

INSTRUMENT REGISTERED TO: KENNETH CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRUNER,SEAN,R
BREATH TEST OPERATOR AGENCY: Kenneth City PD
ARRESTING OFFICER: BRUNER,SEAN
ARRESTING OFFICER AGENCY: Kenneth City PD
LAST AGENCY INSPECTION: 05/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A3
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:55

SUBJECT'S LAST NAME: TOMAS

DATE/TIME OF TEST: 06/17/2017 22:29:10

Diagnostics Check		OK	22:22
Air Blank		0.000	22:22
Control Test		0.080	22:23
Air Blank		0.000	22:23
Subject Sample	# 1	0.020	22:24
Breath Volume		2.171	22:24
Air Blank		0.000	22:25
Air Blank		0.000	22:27
Subject Sample	# 2	0.019	22:27
Breath Volume		3.632	22:27
Air Blank		0.000	22:28
Control Test		0.080	22:28
Air Blank		0.000	22:29
Diagnostics Check		OK	22:29

Data Download Date/Time: 06/22/2017 21:59

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001649

INSTRUMENT REGISTERED TO: KENNETH CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRUNER,SEAN,R
BREATH TEST OPERATOR AGENCY: Kenneth City PD
ARRESTING OFFICER: BRUNER,SEAN
ARRESTING OFFICER AGENCY: Kenneth City PD
LAST AGENCY INSPECTION: 06/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A7
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 19:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:30

SUBJECT'S LAST NAME: OLIVER

DATE/TIME OF TEST: 06/25/2017 19:57:47

Diagnostics Check		OK	19:54	
Air Blank		0.000	19:55	
Control Test		0.079	19:55	
Air Blank		0.000	19:56	
Subject Sample	# 1		19:56	Subject Test Refused
Breath Volume		0.000	19:56	
Air Blank		0.000	19:56	
Control Test		0.080	19:57	
Air Blank		0.000	19:57	
Diagnostics Check		OK	19:57	

Data Download Date/Time: 07/26/2017 22:30

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001649

INSTRUMENT REGISTERED TO: KENNETH CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRUNER,SEAN,R
BREATH TEST OPERATOR AGENCY: Kenneth City PD
ARRESTING OFFICER: BRUNER,SEAN
ARRESTING OFFICER AGENCY: Kenneth City PD
LAST AGENCY INSPECTION: 06/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A7
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 18:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:14

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 06/27/2017 19:43:39

Diagnostics Check		OK	19:36
Air Blank		0.000	19:37
Control Test		0.078	19:37
Air Blank		0.000	19:37
Subject Sample	# 1	0.105	19:38
Breath Volume		2.082	19:38
Air Blank		0.000	19:39
Air Blank		0.000	19:41
Subject Sample	# 2	0.101	19:41
Breath Volume		1.695	19:41
Air Blank		0.000	19:42
Control Test		0.078	19:42
Air Blank		0.000	19:43
Diagnostics Check		OK	19:43

Data Download Date/Time: 07/26/2017 22:30

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIBEIR0,ELAINE,A
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: MORROW,JACE
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A2
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 08:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:12

SUBJECT'S LAST NAME: RIBEIRO

DATE/TIME OF TEST: 06/10/2017 10:42:44

Diagnostics Check		OK	10:36
Air Blank		0.000	10:36
Control Test		0.077	10:36
Air Blank		0.000	10:37
Subject Sample	# 1	0.000	10:38
Breath Volume		1.828	10:38
Air Blank		0.000	10:38
Air Blank		0.000	10:40
Subject Sample	# 2	0.000	10:41
Breath Volume		1.453	10:41
Air Blank		0.000	10:41
Control Test		0.078	10:42
Air Blank		0.000	10:42
Diagnostics Check		OK	10:42

Data Download Date/Time: 07/04/2017 16:22

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEWITT,AMY,L
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: DEWITT,AMY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A2
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 09:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:27

SUBJECT'S LAST NAME: GEE

DATE/TIME OF TEST: 06/13/2017 10:57:48

Diagnostics Check		OK	10:51
Air Blank		0.000	10:51
Control Test		0.078	10:52
Air Blank		0.000	10:52
Subject Sample	# 1	0.000	10:53
Breath Volume		4.097	10:53
Air Blank		0.000	10:54
Air Blank		0.000	10:56
Subject Sample	# 2	0.000	10:56
Breath Volume		3.308	10:56
Air Blank		0.000	10:56
Control Test		0.079	10:57
Air Blank		0.000	10:57
Diagnostics Check		OK	10:57

Data Download Date/Time: 07/04/2017 16:22

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001654

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FIGUEROA,MARIO,E
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: FIGUEROA,MARIO
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 05/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 22:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: BOVA

DATE/TIME OF TEST: 06/01/2017 23:40:13

Diagnostics Check		OK	23:36	
Air Blank		0.000	23:36	
Control Test		0.081	23:37	
Air Blank		0.000	23:37	
Subject Sample	# 1		23:38	RFI Detect
Breath Volume		0.000	23:38	
Air Blank		0.000	23:39	
Control Test		0.081	23:39	
Air Blank		0.000	23:40	
Diagnostics Check		OK	23:40	

Data Download Date/Time: 06/29/2017 15:18

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001654

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FIGUEROA,MARIO,E
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: FIGUEROA,MARIO
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 05/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 22:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: BOVA

DATE/TIME OF TEST: 06/01/2017 23:46:15

Diagnostics Check		OK	23:43	
Air Blank		0.000	23:43	
Control Test		0.080	23:44	
Air Blank		0.000	23:44	
Subject Sample	# 1		23:44	Subject Test Refused
Breath Volume		0.000	23:44	
Air Blank		0.000	23:45	
Control Test		0.080	23:45	
Air Blank		0.000	23:46	
Diagnostics Check		OK	23:46	

Data Download Date/Time: 06/29/2017 15:18

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001654

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARRIOS,MARYURI,O
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: GONZALEZ,JORGE
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 05/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 09:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:56

SUBJECT'S LAST NAME: FAJARDO

DATE/TIME OF TEST: 06/03/2017 10:23:09

Diagnostics Check		OK	10:16
Air Blank		0.000	10:16
Control Test		0.081	10:17
Air Blank		0.000	10:17
Subject Sample	# 1	0.173	10:18
Breath Volume		1.667	10:18
Air Blank		0.000	10:19
Air Blank		0.000	10:20
Subject Sample	# 2	0.186	10:21
Breath Volume		2.933	10:21
Air Blank		0.000	10:22
Control Test		0.079	10:22
Air Blank		0.000	10:23
Diagnostics Check		OK	10:23

Data Download Date/Time: 06/29/2017 15:18

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001654

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SEGURA,OYUKI,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: SEGURA,OYUKI
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 05/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 04:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:15

SUBJECT'S LAST NAME: WYCHE

DATE/TIME OF TEST: 06/09/2017 05:03:21

Diagnostics Check		OK	04:50	
Air Blank		0.000	04:50	
Control Test		0.082	04:50	
Air Blank		0.000	04:51	
Subject Sample	# 1	0.214	04:52	
Breath Volume		1.210	04:52	
Air Blank		0.000	04:53	
Air Blank		0.000	04:54	
Subject Sample	# 2	0.206	04:57	Volume Not Met
Breath Volume		0.941	04:57	
Air Blank		0.000	04:59	
Air Blank		0.000	05:00	
Subject Sample	# 3	0.211	05:01	
Breath Volume		1.132	05:01	
Air Blank		0.000	05:02	
Control Test		0.077	05:02	
Air Blank		0.000	05:03	
Diagnostics Check		OK	05:03	

Data Download Date/Time: 06/29/2017 15:18

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001654

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SEGURA,OYUKI,M
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: SEGURA,OYUKI
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 05/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 05:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:10

SUBJECT'S LAST NAME: TORRES

DATE/TIME OF TEST: 06/11/2017 06:10:18

Diagnostics Check		OK	06:07	
Air Blank		0.000	06:07	
Control Test		0.081	06:08	
Air Blank		0.000	06:08	
Subject Sample	# 1		06:08	Subject Test Refused
Breath Volume		0.000	06:08	
Air Blank		0.000	06:09	
Control Test		0.081	06:09	
Air Blank		0.000	06:10	
Diagnostics Check		OK	06:10	

Data Download Date/Time: 06/29/2017 15:18

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001654

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,R,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: PEREZ,R
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 05/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 23:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:06

SUBJECT'S LAST NAME: AGUILERA

DATE/TIME OF TEST: 06/17/2017 00:35:54

Diagnostics Check		OK	00:27
Air Blank		0.000	00:27
Control Test		0.083	00:27
Air Blank		0.000	00:28
Subject Sample	# 1	0.155	00:31
Breath Volume		1.312	00:31
Air Blank		0.000	00:32
Air Blank		0.000	00:33
Subject Sample	# 2	0.157	00:34
Breath Volume		1.222	00:34
Air Blank		0.000	00:34
Control Test		0.080	00:35
Air Blank		0.000	00:35
Diagnostics Check		OK	00:35

Data Download Date/Time: 06/29/2017 15:18

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001654

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,ANGEL,A
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: HERNANDEZ,ANGEL
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 05/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 21:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: MONCADA

DATE/TIME OF TEST: 06/17/2017 22:33:41

Diagnostics Check	OK	22:32	
Air Blank	0.000	22:32	
Control Test	0.082	22:33	
Air Blank		22:33	RFI Detect
Air Blank	0.000	22:33	

Data Download Date/Time: 06/29/2017 15:18

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001654

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,ANGEL,A
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: HERNANDEZ,ANGEL
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 05/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 22:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:34

SUBJECT'S LAST NAME: MONCADA

DATE/TIME OF TEST: 06/17/2017 22:59:24

Diagnostics Check		OK	22:54	
Air Blank		0.000	22:55	
Control Test		0.082	22:55	
Air Blank		0.000	22:56	
Subject Sample	# 1		22:58	Subject Test Refused
Breath Volume		0.000	22:58	
Air Blank		0.000	22:58	
Control Test		0.082	22:58	
Air Blank		0.000	22:59	
Diagnostics Check		OK	22:59	

Data Download Date/Time: 06/29/2017 15:18

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001654

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARREOLA,RAMIRO,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: ARREOLA,RAMIRO
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 05/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 15:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:00

SUBJECT'S LAST NAME: TUFF

DATE/TIME OF TEST: 06/22/2017 16:22:00

Diagnostics Check	OK	16:20	
Air Blank	0.000	16:20	
Control Test	0.182	16:21	Control Outside Tolerance
Air Blank	0.000	16:22	

Data Download Date/Time: 06/29/2017 15:18

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001654

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARREOLA,RAMIRO,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: ARREOLA,RAMIRO
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 05/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 15:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:00

SUBJECT'S LAST NAME: TUFF

DATE/TIME OF TEST: 06/22/2017 16:28:58

Diagnostics Check	OK	16:28	
Air Blank		16:28	Ambient Fail
Air Blank	0.000	16:28	

Data Download Date/Time: 06/29/2017 15:18

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001654

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARREOLA,RAMIRO,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: ARREOLA,RAMIRO
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 05/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 15:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:00

SUBJECT'S LAST NAME: TUFF

DATE/TIME OF TEST: 06/22/2017 16:34:15

Diagnostics Check	OK	16:33	
Air Blank		16:33	Ambient Fail
Air Blank	0.000	16:34	

Data Download Date/Time: 06/29/2017 15:18

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001654

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARREOLA,RAMIRO,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: ARREOLA,RAMIRO
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 05/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 15:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:00

SUBJECT'S LAST NAME: TUFF

DATE/TIME OF TEST: 06/22/2017 16:38:12

Diagnostics Check	OK	16:37	
Air Blank		16:37	Ambient Fail
Air Blank	0.000	16:38	

Data Download Date/Time: 06/29/2017 15:18

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001654

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARREOLA,RAMIRO,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: ARREOLA,RAMIRO
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 05/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 15:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:00

SUBJECT'S LAST NAME: TUFF

DATE/TIME OF TEST: 06/22/2017 16:43:53

Diagnostics Check	OK	16:42	
Air Blank		16:43	Ambient Fail
Air Blank	0.000	16:43	

Data Download Date/Time: 06/29/2017 15:18

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001654

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARREOLA,RAMIRO,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: ARREOLA,RAMIRO
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 05/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 15:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:00

SUBJECT'S LAST NAME: TUFF

DATE/TIME OF TEST: 06/22/2017 16:54:57

Diagnostics Check	OK	16:53	
Air Blank	0.000	16:53	
Control Test	0.255	16:54	Control Outside Tolerance
Air Blank	0.000	16:54	

Data Download Date/Time: 06/29/2017 15:18

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEPHENS,ZACHARY,A
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: STEPHENS,ZACHARY
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 05/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 640976
DRY GAS EXPIRY DATE: 11/21/2019
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:47

SUBJECT'S LAST NAME: YOUNG

DATE/TIME OF TEST: 06/09/2017 04:21:53

Diagnostics Check		OK	04:09	
Air Blank		0.000	04:09	
Control Test		0.081	04:09	
Air Blank		0.000	04:10	
Subject Sample	# 1	0.184	04:13	Volume Not Met
Breath Volume		0.207	04:13	
Air Blank		0.000	04:14	
Air Blank		0.000	04:15	
Subject Sample	# 2	0.182	04:16	No .020 agreement
Breath Volume		1.320	04:16	
Air Blank		0.000	04:17	
Air Blank		0.000	04:18	
Subject Sample	# 3	0.161	04:20	No .020 agreement
Breath Volume		1.710	04:20	
Air Blank		0.000	04:20	
Control Test		0.080	04:21	
Air Blank		0.000	04:21	
Diagnostics Check		OK	04:21	

Data Download Date/Time: 06/27/2017 14:17

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PATE,MICHAEL,J
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: BRIGGS,JASON
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 06/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 640976
DRY GAS EXPIRY DATE: 11/21/2019
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:08

SUBJECT'S LAST NAME: KETCHERSID II

DATE/TIME OF TEST: 06/28/2017 02:35:35

Diagnostics Check		OK	02:28
Air Blank		0.000	02:29
Control Test		0.082	02:29
Air Blank		0.000	02:30
Subject Sample	# 1	0.211	02:30
Breath Volume		1.304	02:30
Air Blank		0.000	02:31
Air Blank		0.000	02:32
Subject Sample	# 2	0.214	02:34
Breath Volume		2.023	02:34
Air Blank		0.000	02:34
Control Test		0.081	02:35
Air Blank		0.000	02:35
Diagnostics Check		OK	02:35

Data Download Date/Time: 07/23/2017 09:50

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 06/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 640976
DRY GAS EXPIRY DATE: 11/21/2019
VIOLATION CODE: DUI
ARREST TIME: 03:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 06/28/2017 04:58:27

Diagnostics Check		OK	04:51
Air Blank		0.000	04:51
Control Test		0.079	04:52
Air Blank		0.000	04:52
Subject Sample	# 1	0.111	04:53
Breath Volume		2.492	04:53
Air Blank		0.000	04:54
Air Blank		0.000	04:56
Subject Sample	# 2	0.110	04:56
Breath Volume		1.796	04:56
Air Blank		0.000	04:57
Control Test		0.080	04:57
Air Blank		0.000	04:58
Diagnostics Check		OK	04:58

Data Download Date/Time: 07/23/2017 09:50

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001656

INSTRUMENT REGISTERED TO: NORTH MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PINEDA,MIGUEL,A
BREATH TEST OPERATOR AGENCY: North Miami PD
ARRESTING OFFICER: PINEDA,MIGUEL
ARRESTING OFFICER AGENCY: North Miami PD
LAST AGENCY INSPECTION: 05/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:28

SUBJECT'S LAST NAME: SANCHEZ

DATE/TIME OF TEST: 06/25/2017 03:59:05

Diagnostics Check		OK	03:55	
Air Blank		0.000	03:56	
Control Test		0.079	03:56	
Air Blank		0.000	03:57	
Subject Sample	# 1		03:57	Subject Test Refused
Breath Volume		0.000	03:57	
Air Blank		0.000	03:58	
Control Test		0.079	03:58	
Air Blank		0.000	03:58	
Diagnostics Check		OK	03:59	

Data Download Date/Time: 06/28/2017 17:27

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001656

INSTRUMENT REGISTERED TO: NORTH MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PINEDA,MIGUEL,A
BREATH TEST OPERATOR AGENCY: North Miami PD
ARRESTING OFFICER: PINEDA,MIGUEL
ARRESTING OFFICER AGENCY: North Miami PD
LAST AGENCY INSPECTION: 05/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: VILLIERS

DATE/TIME OF TEST: 06/26/2017 03:03:01

Diagnostics Check		OK	02:55
Air Blank		0.000	02:55
Control Test		0.080	02:56
Air Blank		0.000	02:56
Subject Sample	# 1	0.154	02:58
Breath Volume		3.496	02:58
Air Blank		0.000	02:59
Air Blank		0.000	03:00
Subject Sample	# 2	0.156	03:01
Breath Volume		3.257	03:01
Air Blank		0.000	03:02
Control Test		0.079	03:02
Air Blank		0.000	03:02
Diagnostics Check		OK	03:03

Data Download Date/Time: 06/28/2017 17:27

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.