

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: FREY,ROBERT,P

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: ROLPH,A

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 05/26/2017

Date or Time Changed: No

DRY GAS STANDARD LOT #: 26316080A2

DRY GAS EXPIRY DATE: 11/05/2018

VIOLATION CODE: DUI

ARREST TIME: 02:28

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: DEWITT

DATE/TIME OF TEST: 06/02/2017 03:34:07

Diagnostics Check	OK	03:28
Air Blank	0.000	03:28
Control Test	0.080	03:28
Air Blank	0.000	03:29
Subject Sample # 1	0.000	03:29
Breath Volume	2.292	03:29
Air Blank	0.000	03:30
Air Blank	0.000	03:32
Subject Sample # 2	0.000	03:32
Breath Volume	2.769	03:32
Air Blank	0.000	03:33
Control Test	0.080	03:33
Air Blank	0.000	03:34
Diagnostics Check	OK	03:34

Data Download Date/Time: 06/23/2017 17:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: FREY,ROBERT,P

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: ROLPH,A

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 05/26/2017

Date or Time Changed: No

DRY GAS STANDARD LOT #: 26316080A2

DRY GAS EXPIRY DATE: 11/05/2018

VIOLATION CODE: DUI

ARREST TIME: 23:23

GENDER: Female

OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: KOHLES

DATE/TIME OF TEST: 06/06/2017 00:48:50

Diagnostics Check	OK	00:41
Air Blank	0.000	00:42
Control Test	0.080	00:42
Air Blank	0.000	00:42
Subject Sample # 1	0.167	00:44
Breath Volume	2.125	00:44
Air Blank	0.000	00:45
Air Blank	0.000	00:46
Subject Sample # 2	0.162	00:47
Breath Volume	1.496	00:47
Air Blank	0.000	00:47
Control Test	0.079	00:48
Air Blank	0.000	00:48
Diagnostics Check	OK	00:48

Data Download Date/Time: 06/23/2017 17:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIDGEON,ANDREW,W
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: PIDGEON,ANDREW
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 05/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 08:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:37

SUBJECT'S LAST NAME: HENRICK

DATE/TIME OF TEST: 06/10/2017 09:31:46

Diagnostics Check	OK	09:24
Air Blank	0.000	09:25
Control Test	0.080	09:25
Air Blank	0.000	09:26
Subject Sample # 1	0.185	09:26
Breath Volume	1.195	09:26
Air Blank	0.000	09:27
Air Blank	0.000	09:29
Subject Sample # 2	0.171	09:30
Breath Volume	1.476	09:30
Air Blank	0.000	09:30
Control Test	0.079	09:31
Air Blank	0.000	09:31
Diagnostics Check	OK	09:31

Data Download Date/Time: 06/23/2017 17:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: WILLIAMS,B,

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: ROLPH,A

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 05/26/2017

Date or Time Changed: No

DRY GAS STANDARD LOT #: 26316080A2

DRY GAS EXPIRY DATE: 11/05/2018

VIOLATION CODE: DUI

ARREST TIME: 23:21

GENDER: Male

OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: PANTOJA

DATE/TIME OF TEST: 06/11/2017 00:34:23

Diagnostics Check	OK	00:27
Air Blank	0.000	00:28
Control Test	0.081	00:28
Air Blank	0.000	00:28
Subject Sample # 1	0.000	00:29
Breath Volume	3.121	00:29
Air Blank	0.000	00:30
Air Blank	0.000	00:32
Subject Sample # 2	0.000	00:33
Breath Volume	3.429	00:33
Air Blank	0.000	00:33
Control Test	0.079	00:33
Air Blank	0.000	00:34
Diagnostics Check	OK	00:34

Data Download Date/Time: 06/23/2017 17:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: WILLIAMS,B,

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: HOLLCROFT,D

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 05/26/2017

Date or Time Changed: No

DRY GAS STANDARD LOT #: 26316080A2

DRY GAS EXPIRY DATE: 11/05/2018

VIOLATION CODE: DUI

ARREST TIME: 00:33

GENDER: Male

OBSERVATION PERIOD BEGAN: 00:42

SUBJECT'S LAST NAME: SAYMAN

DATE/TIME OF TEST: 06/11/2017 01:09:49

Diagnostics Check	OK	01:03
Air Blank	0.000	01:03
Control Test	0.080	01:04
Air Blank	0.000	01:04
Subject Sample # 1	0.201	01:05
Breath Volume	3.757	01:05
Air Blank	0.000	01:05
Air Blank	0.000	01:07
Subject Sample # 2	0.197	01:08
Breath Volume	3.773	01:08
Air Blank	0.000	01:08
Control Test	0.079	01:09
Air Blank	0.000	01:09
Diagnostics Check	OK	01:09

Data Download Date/Time: 06/23/2017 17:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HAMILTON,J,W

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: HAMILTON,J

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 05/26/2017

Date or Time Changed: No

DRY GAS STANDARD LOT #: 26316080A2

DRY GAS EXPIRY DATE: 11/05/2018

VIOLATION CODE: DUI

ARREST TIME: 03:29

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:29

SUBJECT'S LAST NAME: GRAHAM

DATE/TIME OF TEST: 06/13/2017 04:21:34

Diagnostics Check	OK	04:09	
Air Blank	0.000	04:10	
Control Test	0.080	04:10	
Air Blank	0.000	04:11	
Subject Sample	# 1 0.157	04:14	Volume Not Met
Breath Volume	0.062	04:14	
Air Blank	0.000	04:15	
Air Blank	0.000	04:16	
Subject Sample	# 2 0.153	04:19	Volume Not Met
Breath Volume	0.082	04:19	
Air Blank	0.000	04:20	
Control Test	0.077	04:20	
Air Blank	0.000	04:21	
Diagnostics Check	OK	04:21	

Data Download Date/Time: 06/23/2017 17:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HAMILTON,J,W

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: HAMILTON,J

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 05/26/2017

Date or Time Changed: No

DRY GAS STANDARD LOT #: 26316080A2

DRY GAS EXPIRY DATE: 11/05/2018

VIOLATION CODE: DUI

ARREST TIME: 03:29

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:29

SUBJECT'S LAST NAME: GRAHAM

DATE/TIME OF TEST: 06/13/2017 04:33:14

Diagnostics Check	OK	04:24	
Air Blank	0.000	04:25	
Control Test	0.077	04:25	
Air Blank	0.000	04:25	
Subject Sample	# 1 0.146	04:29	Volume Not Met
Breath Volume	0.070	04:29	
Air Blank	0.000	04:29	
Air Blank	0.000	04:31	
Subject Sample	# 2	04:31	Subject Test Refused
Breath Volume	0.000	04:31	
Air Blank	0.000	04:32	
Control Test	0.077	04:32	
Air Blank	0.000	04:33	
Diagnostics Check	OK	04:33	

Data Download Date/Time: 06/23/2017 17:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LUTZ,EDWIN,J

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: LUTZ,EDWIN

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 05/26/2017

Date or Time Changed: No

DRY GAS STANDARD LOT #: 26316080A2

DRY GAS EXPIRY DATE: 11/05/2018

VIOLATION CODE: DUI

ARREST TIME: 20:58

GENDER: Male

OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: HARDEN

DATE/TIME OF TEST: 06/13/2017 21:34:54

Diagnostics Check	OK	21:27
Air Blank	0.000	21:28
Control Test	0.080	21:28
Air Blank	0.000	21:29
Subject Sample # 1	0.113	21:30
Breath Volume	1.707	21:30
Air Blank	0.000	21:30
Air Blank	0.000	21:32
Subject Sample # 2	0.105	21:33
Breath Volume	1.667	21:33
Air Blank	0.000	21:33
Control Test	0.079	21:34
Air Blank	0.000	21:34
Diagnostics Check	OK	21:34

Data Download Date/Time: 06/23/2017 17:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LUTZ,EDWIN,J

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: LUTZ,EDWIN

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 05/26/2017

Date or Time Changed: No

DRY GAS STANDARD LOT #: 26316080A2

DRY GAS EXPIRY DATE: 11/05/2018

VIOLATION CODE: DUI

ARREST TIME: 21:24

GENDER: Male

OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: SULLIVAN

DATE/TIME OF TEST: 06/14/2017 22:09:52

Diagnostics Check	OK	22:03
Air Blank	0.000	22:04
Control Test	0.080	22:04
Air Blank	0.000	22:04
Subject Sample # 1	0.019	22:05
Breath Volume	1.535	22:05
Air Blank	0.000	22:06
Air Blank	0.000	22:08
Subject Sample # 2	0.020	22:08
Breath Volume	1.484	22:08
Air Blank	0.000	22:08
Control Test	0.080	22:09
Air Blank	0.000	22:09
Diagnostics Check	OK	22:09

Data Download Date/Time: 06/23/2017 17:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LUTZ,EDWIN,J

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: LUTZ,EDWIN

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 05/26/2017

Date or Time Changed: No

DRY GAS STANDARD LOT #: 26316080A2

DRY GAS EXPIRY DATE: 11/05/2018

VIOLATION CODE: DUI

ARREST TIME: 23:42

GENDER: Male

OBSERVATION PERIOD BEGAN: 23:34

SUBJECT'S LAST NAME: KOEHL

DATE/TIME OF TEST: 06/15/2017 00:10:33

Diagnostics Check	OK	00:04
Air Blank	0.000	00:04
Control Test	0.080	00:05
Air Blank	0.000	00:05
Subject Sample # 1	0.124	00:06
Breath Volume	2.640	00:06
Air Blank	0.000	00:06
Air Blank	0.000	00:08
Subject Sample # 2	0.113	00:08
Breath Volume	2.402	00:08
Air Blank	0.000	00:09
Control Test	0.080	00:09
Air Blank	0.000	00:10
Diagnostics Check	OK	00:10

Data Download Date/Time: 06/23/2017 17:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: FREY,ROBERT,P

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: KING,P

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 05/26/2017

Date or Time Changed: No

DRY GAS STANDARD LOT #: 26316080A2

DRY GAS EXPIRY DATE: 11/05/2018

VIOLATION CODE: DUI

ARREST TIME: 21:54

GENDER: Male

OBSERVATION PERIOD BEGAN: 22:55

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 06/18/2017 23:33:48

Diagnostics Check	OK	23:27
Air Blank	0.000	23:28
Control Test	0.080	23:28
Air Blank	0.000	23:28
Subject Sample # 1	0.119	23:29
Breath Volume	1.367	23:29
Air Blank	0.000	23:30
Air Blank	0.000	23:31
Subject Sample # 2	0.122	23:32
Breath Volume	1.191	23:32
Air Blank	0.000	23:32
Control Test	0.079	23:33
Air Blank	0.000	23:33
Diagnostics Check	OK	23:33

Data Download Date/Time: 06/23/2017 17:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: FREY,ROBERT,P

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: BELL,GORDON

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 05/26/2017

Date or Time Changed: No

DRY GAS STANDARD LOT #: 26316080A2

DRY GAS EXPIRY DATE: 11/05/2018

VIOLATION CODE: DUI

ARREST TIME: 03:41

GENDER: Female

OBSERVATION PERIOD BEGAN: 04:47

SUBJECT'S LAST NAME: RYAN

DATE/TIME OF TEST: 06/19/2017 05:19:14

Diagnostics Check	OK	05:12
Air Blank	0.000	05:13
Control Test	0.080	05:13
Air Blank	0.000	05:14
Subject Sample # 1	0.206	05:14
Breath Volume	1.699	05:14
Air Blank	0.000	05:15
Air Blank	0.000	05:17
Subject Sample # 2	0.198	05:17
Breath Volume	1.640	05:17
Air Blank	0.000	05:18
Control Test	0.080	05:18
Air Blank	0.000	05:19
Diagnostics Check	OK	05:19

Data Download Date/Time: 06/23/2017 17:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HAMILTON,J,W

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: HAMILTON,J

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 05/26/2017

Date or Time Changed: No

DRY GAS STANDARD LOT #: 26316080A2

DRY GAS EXPIRY DATE: 11/05/2018

VIOLATION CODE: DUI

ARREST TIME: 23:25

GENDER: Female

OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: HENRY

DATE/TIME OF TEST: 06/19/2017 23:59:29

Diagnostics Check	OK	23:51
Air Blank	0.000	23:52
Control Test	0.079	23:52
Air Blank	0.000	23:52
Subject Sample # 1	0.186	23:54
Breath Volume	1.386	23:54
Air Blank	0.000	23:55
Air Blank	0.000	23:57
Subject Sample # 2	0.179	23:57
Breath Volume	1.273	23:57
Air Blank	0.000	23:58
Control Test	0.078	23:58
Air Blank	0.000	23:59
Diagnostics Check	OK	23:59

Data Download Date/Time: 06/23/2017 17:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LUTZ,EDWIN,J

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: LUTZ,EDWIN

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 05/26/2017

Date or Time Changed: No

DRY GAS STANDARD LOT #: 26316080A2

DRY GAS EXPIRY DATE: 11/05/2018

VIOLATION CODE: DUI

ARREST TIME: 20:52

GENDER: Male

OBSERVATION PERIOD BEGAN: 21:06

SUBJECT'S LAST NAME: WATSON

DATE/TIME OF TEST: 06/20/2017 21:33:25

Diagnostics Check	OK	21:26
Air Blank	0.000	21:27
Control Test	0.079	21:27
Air Blank	0.000	21:28
Subject Sample # 1	0.280	21:28
Breath Volume	2.167	21:28
Air Blank	0.000	21:29
Air Blank	0.000	21:31
Subject Sample # 2	0.286	21:31
Breath Volume	2.277	21:31
Air Blank	0.000	21:32
Control Test	0.078	21:32
Air Blank	0.000	21:33
Diagnostics Check	OK	21:33

Data Download Date/Time: 06/23/2017 17:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ENGLAND,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: MEDINA,JOSE
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 05/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 16:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:58

SUBJECT'S LAST NAME: CAGNI

DATE/TIME OF TEST: 06/02/2017 17:38:07

Diagnostics Check	OK	17:31
Air Blank	0.000	17:31
Control Test	0.080	17:32
Air Blank	0.000	17:32
Subject Sample # 1	0.268	17:33
Breath Volume	2.429	17:33
Air Blank	0.000	17:34
Air Blank	0.000	17:35
Subject Sample # 2	0.278	17:36
Breath Volume	1.695	17:36
Air Blank	0.000	17:37
Control Test	0.079	17:37
Air Blank	0.000	17:38
Diagnostics Check	OK	17:38

Data Download Date/Time: 06/22/2017 15:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EISENHARDT,ERIC,J
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: EISENHARDT,ERIC
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 05/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 06/06/2017 22:32:27

Diagnostics Check	OK	22:26	
Air Blank	0.000	22:26	
Control Test	0.081	22:26	
Air Blank	0.000	22:27	
Subject Sample	# 1 0.221	22:28	
Breath Volume	1.386	22:28	
Air Blank	0.000	22:28	
Air Blank	0.000	22:30	
Subject Sample	# 2	22:31	RFI Detect
Breath Volume	0.000	22:31	
Air Blank	0.000	22:31	
Control Test	0.080	22:31	
Air Blank	0.000	22:32	
Diagnostics Check	OK	22:32	

Data Download Date/Time: 06/22/2017 15:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EISENHARDT,ERIC,J
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: EISENHARDT,ERIC
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 05/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 06/06/2017 22:41:37

Diagnostics Check	OK	22:34
Air Blank	0.000	22:35
Control Test	0.080	22:35
Air Blank	0.000	22:36
Subject Sample # 1	0.231	22:36
Breath Volume	1.335	22:36
Air Blank	0.000	22:37
Air Blank	0.000	22:39
Subject Sample # 2	0.224	22:39
Breath Volume	1.312	22:39
Air Blank	0.000	22:40
Control Test	0.080	22:40
Air Blank	0.000	22:41
Diagnostics Check	OK	22:41

Data Download Date/Time: 06/22/2017 15:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ENGLAND,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: ENGLAND,CHRISTOPHER
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 05/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:10

SUBJECT'S LAST NAME: ARDITI

DATE/TIME OF TEST: 06/08/2017 21:38:21

Diagnostics Check	OK	21:31
Air Blank	0.000	21:31
Control Test	0.081	21:31
Air Blank	0.000	21:32
Subject Sample # 1	0.165	21:33
Breath Volume	2.757	21:33
Air Blank	0.000	21:34
Air Blank	0.000	21:36
Subject Sample # 2	0.154	21:36
Breath Volume	1.867	21:36
Air Blank	0.000	21:37
Control Test	0.080	21:37
Air Blank	0.000	21:38
Diagnostics Check	OK	21:38

Data Download Date/Time: 06/22/2017 15:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NEGERSMITH,D,M
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: NEGERSMITH,D
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 05/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:58

SUBJECT'S LAST NAME: LODATO

DATE/TIME OF TEST: 06/09/2017 01:44:56

Diagnostics Check	OK	01:33	
Air Blank	0.000	01:33	
Control Test	0.080	01:34	
Air Blank	0.000	01:34	
Subject Sample	# 1	01:37	No Sample Provided
Breath Volume	0.000	01:37	
Air Blank	0.000	01:38	
Air Blank	0.000	01:40	
Subject Sample	# 2	01:43	No Sample Provided
Breath Volume	0.000	01:43	
Air Blank	0.000	01:44	
Control Test	0.081	01:44	
Air Blank	0.000	01:44	
Diagnostics Check	OK	01:44	

Data Download Date/Time: 06/22/2017 15:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NUGENT,DAVID,A
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: NUGENT,DAVID
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 05/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:09

SUBJECT'S LAST NAME: HERNANDEZ CARMONA

DATE/TIME OF TEST: 06/09/2017 22:49:38

Diagnostics Check	OK	22:38	
Air Blank	0.000	22:38	
Control Test	0.080	22:38	
Air Blank	0.000	22:39	
Subject Sample	# 1 0.176	22:42	Volume Not Met
Breath Volume	0.027	22:42	
Air Blank	0.000	22:43	
Air Blank	0.000	22:45	
Subject Sample	# 2	22:48	No Sample Provided
Breath Volume	0.000	22:48	
Air Blank	0.000	22:48	
Control Test	0.078	22:49	
Air Blank	0.000	22:49	
Diagnostics Check	OK	22:49	

Data Download Date/Time: 06/22/2017 15:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NEGERSMITH,D,M
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: NEGERSMITH,D
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 05/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: BURNS

DATE/TIME OF TEST: 06/11/2017 01:00:29

Diagnostics Check	OK	00:53
Air Blank	0.000	00:53
Control Test	0.081	00:54
Air Blank	0.000	00:54
Subject Sample # 1	0.136	00:55
Breath Volume	1.585	00:55
Air Blank	0.000	00:56
Air Blank	0.000	00:58
Subject Sample # 2	0.131	00:58
Breath Volume	2.195	00:58
Air Blank	0.000	00:59
Control Test	0.079	00:59
Air Blank	0.000	01:00
Diagnostics Check	OK	01:00

Data Download Date/Time: 06/22/2017 15:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NUGENT,DAVID,A
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: NUGENT,DAVID
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 05/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 15:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:30

SUBJECT'S LAST NAME: DE LA CRUZ-HERNANDEZ

DATE/TIME OF TEST: 06/16/2017 16:07:25

Diagnostics Check	OK	15:53	
Air Blank	0.000	15:53	
Control Test	0.081	15:54	
Air Blank	0.000	15:54	
Subject Sample	# 1 0.223	15:57	Volume Not Met
Breath Volume	0.343	15:57	
Air Blank	0.000	15:58	
Air Blank	0.000	16:00	
Subject Sample	# 2 0.239	16:01	
Breath Volume	1.640	16:01	
Air Blank	0.000	16:02	
Air Blank	0.000	16:03	
Subject Sample	# 3 0.242	16:05	
Breath Volume	1.789	16:05	
Air Blank	0.000	16:06	
Control Test	0.079	16:06	
Air Blank	0.000	16:07	
Diagnostics Check	OK	16:07	

Data Download Date/Time: 06/22/2017 15:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NUGENT,DAVID,A
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: NUGENT,DAVID
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 05/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:56

SUBJECT'S LAST NAME: SIMPKINS

DATE/TIME OF TEST: 06/16/2017 22:24:32

Diagnostics Check	OK	22:17
Air Blank	0.000	22:17
Control Test	0.080	22:17
Air Blank	0.000	22:18
Subject Sample # 1	0.105	22:19
Breath Volume	2.175	22:19
Air Blank	0.000	22:19
Air Blank	0.000	22:21
Subject Sample # 2	0.102	22:22
Breath Volume	1.785	22:22
Air Blank	0.000	22:23
Control Test	0.081	22:23
Air Blank	0.000	22:24
Diagnostics Check	OK	22:24

Data Download Date/Time: 06/22/2017 15:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NEGERSMITH,D,M
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: NEGERSMITH,D
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 05/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:07

SUBJECT'S LAST NAME: BECKER-SHEEHAN

DATE/TIME OF TEST: 06/17/2017 00:41:06

Diagnostics Check	OK	00:34
Air Blank	0.000	00:35
Control Test	0.081	00:35
Air Blank	0.000	00:36
Subject Sample # 1	0.000	00:36
Breath Volume	2.121	00:36
Air Blank	0.000	00:37
Air Blank	0.000	00:39
Subject Sample # 2	0.000	00:39
Breath Volume	1.457	00:39
Air Blank	0.000	00:40
Control Test	0.080	00:40
Air Blank	0.000	00:41
Diagnostics Check	OK	00:41

Data Download Date/Time: 06/22/2017 15:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EISENHARDT,ERIC,J
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: EISENHARDT,ERIC
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 05/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: BETTS

DATE/TIME OF TEST: 06/21/2017 20:27:33

Diagnostics Check	OK	20:21
Air Blank	0.000	20:21
Control Test	0.080	20:22
Air Blank	0.000	20:22
Subject Sample # 1	0.000	20:23
Breath Volume	2.636	20:23
Air Blank	0.000	20:23
Air Blank	0.000	20:25
Subject Sample # 2	0.000	20:26
Breath Volume	2.066	20:26
Air Blank	0.000	20:26
Control Test	0.080	20:26
Air Blank	0.000	20:27
Diagnostics Check	OK	20:27

Data Download Date/Time: 06/22/2017 15:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NEGERSMITH,D,M
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: NEGERSMITH,D
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 05/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: MARLEY

DATE/TIME OF TEST: 06/22/2017 03:11:46

Diagnostics Check	OK	03:04
Air Blank	0.000	03:05
Control Test	0.081	03:05
Air Blank	0.000	03:06
Subject Sample # 1	0.154	03:06
Breath Volume	1.632	03:06
Air Blank	0.000	03:07
Air Blank	0.000	03:09
Subject Sample # 2	0.152	03:10
Breath Volume	1.394	03:10
Air Blank	0.000	03:10
Control Test	0.080	03:11
Air Blank	0.000	03:11
Diagnostics Check	OK	03:11

Data Download Date/Time: 06/22/2017 15:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBER,JOSEPH,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MOELLER,R.H.
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: CLARK

DATE/TIME OF TEST: 06/01/2017 03:26:14

Diagnostics Check	OK	03:20
Air Blank	0.000	03:20
Control Test	0.079	03:20
Air Blank	0.000	03:21
Subject Sample # 1	0.198	03:21
Breath Volume	1.980	03:21
Air Blank	0.000	03:22
Air Blank	0.000	03:24
Subject Sample # 2	0.196	03:24
Breath Volume	2.285	03:24
Air Blank	0.000	03:25
Control Test	0.080	03:25
Air Blank	0.000	03:26
Diagnostics Check	OK	03:26

Data Download Date/Time: 06/15/2017 12:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON, BRADEN, J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BYRNE, J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: TAVARES

DATE/TIME OF TEST: 06/02/2017 03:01:17

Diagnostics Check	OK	02:55
Air Blank	0.000	02:55
Control Test	0.080	02:55
Air Blank	0.000	02:56
Subject Sample # 1	0.066	02:56
Breath Volume	3.550	02:56
Air Blank	0.000	02:57
Air Blank	0.000	02:59
Subject Sample # 2	0.065	02:59
Breath Volume	3.570	02:59
Air Blank	0.000	03:00
Control Test	0.079	03:00
Air Blank	0.000	03:01
Diagnostics Check	OK	03:01

Data Download Date/Time: 06/15/2017 12:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBER,JOSEPH,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: EANES,J.L.
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:12

SUBJECT'S LAST NAME: EVANS

DATE/TIME OF TEST: 06/02/2017 04:38:44

Diagnostics Check	OK	04:32
Air Blank	0.000	04:32
Control Test	0.080	04:33
Air Blank	0.000	04:33
Subject Sample # 1	0.137	04:34
Breath Volume	1.308	04:34
Air Blank	0.000	04:34
Air Blank	0.000	04:36
Subject Sample # 2	0.135	04:37
Breath Volume	1.558	04:37
Air Blank	0.000	04:37
Control Test	0.079	04:38
Air Blank	0.000	04:38
Diagnostics Check	OK	04:38

Data Download Date/Time: 06/15/2017 12:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON, BRADEN, J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BROWN, T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:51

SUBJECT'S LAST NAME: BADILLO

DATE/TIME OF TEST: 06/03/2017 01:18:55

Diagnostics Check	OK	01:12
Air Blank	0.000	01:12
Control Test	0.080	01:12
Air Blank	0.000	01:13
Subject Sample # 1	0.119	01:14
Breath Volume	2.871	01:14
Air Blank	0.000	01:14
Air Blank	0.000	01:16
Subject Sample # 2	0.122	01:17
Breath Volume	3.125	01:17
Air Blank	0.000	01:17
Control Test	0.079	01:18
Air Blank	0.000	01:18
Diagnostics Check	OK	01:18

Data Download Date/Time: 06/15/2017 12:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURNS,D,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BYRNE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:11

SUBJECT'S LAST NAME: SOOKNANDAN

DATE/TIME OF TEST: 06/03/2017 02:35:24

Diagnostics Check	OK	02:31	
Air Blank	0.000	02:32	
Control Test	0.080	02:32	
Air Blank	0.000	02:33	
Subject Sample	# 1	02:34	Subject Test Refused
Breath Volume	0.000	02:34	
Air Blank	0.000	02:34	
Control Test	0.079	02:34	
Air Blank	0.000	02:35	
Diagnostics Check	OK	02:35	

Data Download Date/Time: 06/15/2017 12:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORTAGUS,B,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: GRIFFIS,M
ARRESTING OFFICER AGENCY: FFWCC (Lake City)
LAST AGENCY INSPECTION: 05/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: BUI
ARREST TIME: 16:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:14

SUBJECT'S LAST NAME: WINNON

DATE/TIME OF TEST: 06/03/2017 19:41:36

Diagnostics Check	OK	19:34
Air Blank	0.000	19:35
Control Test	0.079	19:35
Air Blank	0.000	19:36
Subject Sample # 1	0.164	19:36
Breath Volume	2.007	19:36
Air Blank	0.000	19:37
Air Blank	0.000	19:39
Subject Sample # 2	0.163	19:40
Breath Volume	1.593	19:40
Air Blank	0.000	19:40
Control Test	0.079	19:40
Air Blank	0.000	19:41
Diagnostics Check	OK	19:41

Data Download Date/Time: 06/15/2017 12:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORTAGUS,B,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: TORRES,W
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 05/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: LEWIS

DATE/TIME OF TEST: 06/03/2017 22:08:35

Diagnostics Check	OK	22:05	
Air Blank	0.000	22:05	
Control Test	0.080	22:06	
Air Blank	0.000	22:06	
Subject Sample	# 1	22:07	Subject Test Refused
Breath Volume	0.000	22:07	
Air Blank	0.000	22:07	
Control Test	0.080	22:07	
Air Blank	0.000	22:08	
Diagnostics Check	OK	22:08	

Data Download Date/Time: 06/15/2017 12:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON, BRADEN, J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BRYANT, R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:07

SUBJECT'S LAST NAME: ARREGUIN-HERNANDEZ

DATE/TIME OF TEST: 06/05/2017 01:33:52

Diagnostics Check	OK	01:27
Air Blank	0.000	01:27
Control Test	0.079	01:28
Air Blank	0.000	01:28
Subject Sample # 1	0.131	01:29
Breath Volume	3.496	01:29
Air Blank	0.000	01:29
Air Blank	0.000	01:31
Subject Sample # 2	0.131	01:32
Breath Volume	3.261	01:32
Air Blank	0.000	01:32
Control Test	0.080	01:33
Air Blank	0.000	01:33
Diagnostics Check	OK	01:33

Data Download Date/Time: 06/15/2017 12:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLARIDA,K,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BRYANT,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:17

SUBJECT'S LAST NAME: STENGEL

DATE/TIME OF TEST: 06/05/2017 21:41:51

Diagnostics Check	OK	21:38	
Air Blank	0.000	21:39	
Control Test	0.080	21:39	
Air Blank	0.000	21:39	
Subject Sample	# 1	21:40	Slope Not Met
Breath Volume	0.890	21:40	
Air Blank	0.000	21:40	
Control Test	0.080	21:41	
Air Blank	0.000	21:41	
Diagnostics Check	OK	21:41	

Data Download Date/Time: 06/15/2017 12:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLARIDA,K,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BRYANT,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:17

SUBJECT'S LAST NAME: STENGEL

DATE/TIME OF TEST: 06/05/2017 21:50:21

Diagnostics Check	OK	21:43
Air Blank	0.000	21:44
Control Test	0.080	21:44
Air Blank	0.000	21:45
Subject Sample # 1	0.195	21:45
Breath Volume	2.636	21:45
Air Blank	0.000	21:46
Air Blank	0.000	21:48
Subject Sample # 2	0.185	21:48
Breath Volume	3.031	21:48
Air Blank	0.000	21:49
Control Test	0.079	21:49
Air Blank	0.000	21:50
Diagnostics Check	OK	21:50

Data Download Date/Time: 06/15/2017 12:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON,BRADEN,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: GIBSON,D
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 05/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: MCNULTY

DATE/TIME OF TEST: 06/06/2017 01:59:21

Diagnostics Check	OK	01:56	
Air Blank	0.000	01:56	
Control Test	0.080	01:57	
Air Blank	0.000	01:57	
Subject Sample	# 1	01:57	Subject Test Refused
Breath Volume	0.000	01:57	
Air Blank	0.000	01:58	
Control Test	0.080	01:58	
Air Blank	0.000	01:59	
Diagnostics Check	OK	01:59	

Data Download Date/Time: 06/15/2017 12:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBER,JOSEPH,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CAMACHO,K.R.
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:47

SUBJECT'S LAST NAME: KANUCKLE

DATE/TIME OF TEST: 06/07/2017 01:15:44

Diagnostics Check	OK	01:09
Air Blank	0.000	01:09
Control Test	0.080	01:10
Air Blank	0.000	01:10
Subject Sample # 1	0.000	01:11
Breath Volume	2.257	01:11
Air Blank	0.000	01:11
Air Blank	0.000	01:13
Subject Sample # 2	0.000	01:14
Breath Volume	2.285	01:14
Air Blank	0.000	01:14
Control Test	0.080	01:15
Air Blank	0.000	01:15
Diagnostics Check	OK	01:15

Data Download Date/Time: 06/15/2017 12:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: RODRIGUEZ,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:41

SUBJECT'S LAST NAME: NEWSOM

DATE/TIME OF TEST: 06/08/2017 04:09:41

Diagnostics Check	OK	04:03
Air Blank	0.000	04:03
Control Test	0.081	04:04
Air Blank	0.000	04:04
Subject Sample # 1	0.060	04:05
Breath Volume	1.500	04:05
Air Blank	0.000	04:05
Air Blank	0.000	04:07
Subject Sample # 2	0.063	04:08
Breath Volume	1.183	04:08
Air Blank	0.000	04:08
Control Test	0.080	04:09
Air Blank	0.000	04:09
Diagnostics Check	OK	04:09

Data Download Date/Time: 06/15/2017 12:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JORDAN,A
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 05/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:15

SUBJECT'S LAST NAME: SOTO

DATE/TIME OF TEST: 06/08/2017 05:39:08

Diagnostics Check	OK	05:36	
Air Blank	0.000	05:36	
Control Test	0.079	05:37	
Air Blank	0.000	05:37	
Subject Sample	# 1	05:37	Subject Test Refused
Breath Volume	0.000	05:37	
Air Blank	0.000	05:38	
Control Test	0.080	05:38	
Air Blank	0.000	05:39	
Diagnostics Check	OK	05:39	

Data Download Date/Time: 06/15/2017 12:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBER,JOSEPH,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FARLEY,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 05/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: EDDY

DATE/TIME OF TEST: 06/08/2017 23:34:05

Diagnostics Check	OK	23:31	
Air Blank	0.000	23:31	
Control Test	0.080	23:32	
Air Blank	0.000	23:32	
Subject Sample	# 1	23:32	Subject Test Refused
Breath Volume	0.000	23:32	
Air Blank	0.000	23:33	
Control Test	0.080	23:33	
Air Blank	0.000	23:34	
Diagnostics Check	OK	23:34	

Data Download Date/Time: 06/15/2017 12:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBER,JOSEPH,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HOOVER,J.K.
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:36

SUBJECT'S LAST NAME: ROBLES

DATE/TIME OF TEST: 06/09/2017 05:02:07

Diagnostics Check	OK	04:58	
Air Blank	0.000	04:59	
Control Test	0.080	04:59	
Air Blank	0.000	05:00	
Subject Sample	# 1	05:00	Subject Test Refused
Breath Volume	0.000	05:00	
Air Blank	0.000	05:01	
Control Test	0.081	05:01	
Air Blank	0.000	05:02	
Diagnostics Check	OK	05:02	

Data Download Date/Time: 06/15/2017 12:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBER,JOSEPH,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHITE,J.M.
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:01

SUBJECT'S LAST NAME: HANSEN

DATE/TIME OF TEST: 06/10/2017 01:28:09

Diagnostics Check	OK	01:21
Air Blank	0.000	01:22
Control Test	0.080	01:22
Air Blank	0.000	01:22
Subject Sample # 1	0.225	01:23
Breath Volume	2.890	01:23
Air Blank	0.000	01:23
Air Blank	0.000	01:25
Subject Sample # 2	0.219	01:26
Breath Volume	3.058	01:26
Air Blank	0.000	01:27
Control Test	0.079	01:27
Air Blank	0.000	01:28
Diagnostics Check	OK	01:28

Data Download Date/Time: 06/15/2017 12:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBER,JOSEPH,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HOOVER,J.K.
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:34

SUBJECT'S LAST NAME: CUMMINGS

DATE/TIME OF TEST: 06/10/2017 03:00:25

Diagnostics Check	OK	02:57	
Air Blank	0.000	02:57	
Control Test	0.079	02:58	
Air Blank	0.000	02:58	
Subject Sample	# 1	02:59	Subject Test Refused
Breath Volume	0.000	02:59	
Air Blank	0.000	02:59	
Control Test	0.080	02:59	
Air Blank	0.000	03:00	
Diagnostics Check	OK	03:00	

Data Download Date/Time: 06/15/2017 12:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: GREEN,J
ARRESTING OFFICER AGENCY: Atlantic Beach PD
LAST AGENCY INSPECTION: 05/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:36

SUBJECT'S LAST NAME: MCDONALD

DATE/TIME OF TEST: 06/10/2017 03:59:19

Diagnostics Check	OK	03:56	
Air Blank	0.000	03:56	
Control Test	0.079	03:57	
Air Blank	0.000	03:57	
Subject Sample	# 1	03:57	Subject Test Refused
Breath Volume	0.000	03:57	
Air Blank	0.000	03:58	
Control Test	0.080	03:58	
Air Blank	0.000	03:59	
Diagnostics Check	OK	03:59	

Data Download Date/Time: 06/15/2017 12:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: RODRIGUEZ,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:44

SUBJECT'S LAST NAME: LARRY

DATE/TIME OF TEST: 06/10/2017 05:11:35

Diagnostics Check	OK	05:08	
Air Blank	0.000	05:09	
Control Test	0.080	05:09	
Air Blank	0.000	05:09	
Subject Sample	# 1	05:10	Subject Test Refused
Breath Volume	0.000	05:10	
Air Blank	0.000	05:10	
Control Test	0.080	05:10	
Air Blank	0.000	05:11	
Diagnostics Check	OK	05:11	

Data Download Date/Time: 06/15/2017 12:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBER,JOSEPH,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MERRERO,J.L.
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:18

SUBJECT'S LAST NAME: LOPEZ LOPEZ

DATE/TIME OF TEST: 06/11/2017 02:45:05

Diagnostics Check	OK	02:38
Air Blank	0.000	02:39
Control Test	0.080	02:39
Air Blank	0.000	02:40
Subject Sample # 1	0.172	02:40
Breath Volume	2.609	02:40
Air Blank	0.000	02:41
Air Blank	0.000	02:42
Subject Sample # 2	0.170	02:43
Breath Volume	2.195	02:43
Air Blank	0.000	02:44
Control Test	0.080	02:44
Air Blank	0.000	02:45
Diagnostics Check	OK	02:45

Data Download Date/Time: 06/15/2017 12:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBER,JOSEPH,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BYRNE,J.F.
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:34

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 06/11/2017 03:57:10

Diagnostics Check	OK	03:55	
Air Blank	0.000	03:55	
Control Test	0.080	03:55	
Air Blank	0.000	03:56	
Subject Sample	# 1	03:56	Subject Test Refused
Breath Volume	0.000	03:56	
Air Blank		03:57	RFI Detect
Air Blank		03:57	RFI Detect

Data Download Date/Time: 06/15/2017 12:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBER,JOSEPH,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MOORE,A.L.
ARRESTING OFFICER AGENCY: Atlantic Beach PD
LAST AGENCY INSPECTION: 05/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:21

SUBJECT'S LAST NAME: IRVIN

DATE/TIME OF TEST: 06/11/2017 04:57:11

Diagnostics Check	OK	04:42	
Air Blank	0.000	04:42	
Control Test	0.080	04:42	
Air Blank	0.000	04:43	
Subject Sample	# 1 0.193	04:46	
Breath Volume	1.902	04:46	
Air Blank	0.000	04:47	
Air Blank	0.000	04:48	
Subject Sample	# 2 0.177	04:52	Slope Not Level
Breath Volume	1.605	04:52	
Air Blank	0.000	04:52	
Air Blank	0.000	04:54	
Subject Sample	# 3 0.190	04:55	
Breath Volume	2.394	04:55	
Air Blank	0.000	04:56	
Control Test	0.077	04:56	
Air Blank	0.000	04:57	
Diagnostics Check	OK	04:57	

Data Download Date/Time: 06/15/2017 12:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBER,JOSEPH,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BYRNE,J.F.
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:39

SUBJECT'S LAST NAME: MARROQUIN SALAS

DATE/TIME OF TEST: 06/12/2017 02:02:41

Diagnostics Check	OK	01:59	
Air Blank	0.000	02:00	
Control Test	0.080	02:00	
Air Blank	0.000	02:00	
Subject Sample	# 1	02:01	Subject Test Refused
Breath Volume	0.000	02:01	
Air Blank	0.000	02:01	
Control Test	0.081	02:02	
Air Blank	0.000	02:02	
Diagnostics Check	OK	02:02	

Data Download Date/Time: 06/15/2017 12:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBER,JOSEPH,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: VANHOF,J.H.
ARRESTING OFFICER AGENCY: Atlantic Beach PD
LAST AGENCY INSPECTION: 05/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: SUBHAM

DATE/TIME OF TEST: 06/12/2017 03:34:12

Diagnostics Check	OK	03:28
Air Blank	0.000	03:28
Control Test	0.079	03:28
Air Blank	0.000	03:29
Subject Sample # 1	0.182	03:29
Breath Volume	1.808	03:29
Air Blank	0.000	03:30
Air Blank	0.000	03:32
Subject Sample # 2	0.185	03:32
Breath Volume	1.769	03:32
Air Blank	0.000	03:33
Control Test	0.080	03:33
Air Blank	0.000	03:34
Diagnostics Check	OK	03:34

Data Download Date/Time: 06/15/2017 12:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBER,JOSEPH,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BAIS,W.
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:22

SUBJECT'S LAST NAME: NICHOLS

DATE/TIME OF TEST: 06/13/2017 02:48:29

Diagnostics Check	OK	02:42
Air Blank	0.000	02:42
Control Test	0.080	02:43
Air Blank	0.000	02:43
Subject Sample # 1	0.207	02:43
Breath Volume	3.320	02:43
Air Blank	0.000	02:44
Air Blank	0.000	02:46
Subject Sample # 2	0.214	02:46
Breath Volume	2.500	02:46
Air Blank	0.000	02:47
Control Test	0.080	02:47
Air Blank	0.000	02:48
Diagnostics Check	OK	02:48

Data Download Date/Time: 06/15/2017 12:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON, BRADEN, J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: STANLEY, T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:54

SUBJECT'S LAST NAME: STURM

DATE/TIME OF TEST: 06/16/2017 00:21:14

Diagnostics Check	OK	00:14
Air Blank	0.000	00:15
Control Test	0.080	00:15
Air Blank	0.000	00:16
Subject Sample # 1	0.000	00:16
Breath Volume	2.488	00:16
Air Blank	0.000	00:17
Air Blank	0.000	00:19
Subject Sample # 2	0.000	00:19
Breath Volume	2.394	00:19
Air Blank	0.000	00:20
Control Test	0.080	00:20
Air Blank	0.000	00:21
Diagnostics Check	OK	00:21

Data Download Date/Time: 06/27/2017 14:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCOTT,J,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: THOMPSON,M
ARRESTING OFFICER AGENCY: Atlantic Beach PD
LAST AGENCY INSPECTION: 06/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:41

SUBJECT'S LAST NAME: SEALS

DATE/TIME OF TEST: 06/16/2017 03:10:30

Diagnostics Check	OK	03:07	
Air Blank	0.000	03:08	
Control Test	0.080	03:08	
Air Blank	0.000	03:08	
Subject Sample	# 1	03:09	Subject Test Refused
Breath Volume	0.000	03:09	
Air Blank	0.000	03:09	
Control Test	0.081	03:09	
Air Blank	0.000	03:10	
Diagnostics Check	OK	03:10	

Data Download Date/Time: 06/27/2017 14:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON, BRADEN, J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DAIGLE, B
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:23

SUBJECT'S LAST NAME: MARSHALL

DATE/TIME OF TEST: 06/16/2017 03:49:27

Diagnostics Check	OK	03:43
Air Blank	0.000	03:43
Control Test	0.081	03:43
Air Blank	0.000	03:44
Subject Sample # 1	0.000	03:45
Breath Volume	3.628	03:45
Air Blank	0.000	03:45
Air Blank	0.000	03:47
Subject Sample # 2	0.000	03:48
Breath Volume	3.621	03:48
Air Blank	0.000	03:48
Control Test	0.080	03:48
Air Blank	0.000	03:49
Diagnostics Check	OK	03:49

Data Download Date/Time: 06/27/2017 14:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIGHTSEY,ROBERT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FARLEY,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 06/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:32

SUBJECT'S LAST NAME: EDWARDS

DATE/TIME OF TEST: 06/16/2017 23:59:24

Diagnostics Check	OK	23:53
Air Blank	0.000	23:53
Control Test	0.080	23:53
Air Blank	0.000	23:54
Subject Sample # 1	0.135	23:55
Breath Volume	1.937	23:55
Air Blank	0.000	23:55
Air Blank	0.000	23:57
Subject Sample # 2	0.133	23:57
Breath Volume	2.066	23:57
Air Blank	0.000	23:58
Control Test	0.079	23:58
Air Blank	0.000	23:59
Diagnostics Check	OK	23:59

Data Download Date/Time: 06/27/2017 14:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIGHTSEY,ROBERT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CAMACHO,K
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 06/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:31

SUBJECT'S LAST NAME: JAKES

DATE/TIME OF TEST: 06/17/2017 00:57:44

Diagnostics Check	OK	00:51
Air Blank	0.000	00:51
Control Test	0.081	00:52
Air Blank	0.000	00:52
Subject Sample # 1	0.123	00:53
Breath Volume	2.410	00:53
Air Blank	0.000	00:53
Air Blank	0.000	00:55
Subject Sample # 2	0.124	00:56
Breath Volume	2.882	00:56
Air Blank	0.000	00:56
Control Test	0.080	00:57
Air Blank	0.000	00:57
Diagnostics Check	OK	00:57

Data Download Date/Time: 06/27/2017 14:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIGHTSEY,ROBERT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HEALY,G
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 06/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:51

SUBJECT'S LAST NAME: CRAWFORD

DATE/TIME OF TEST: 06/17/2017 02:18:33

Diagnostics Check	OK	02:12
Air Blank	0.000	02:12
Control Test	0.079	02:12
Air Blank	0.000	02:13
Subject Sample # 1	0.219	02:13
Breath Volume	3.265	02:13
Air Blank	0.000	02:14
Air Blank	0.000	02:16
Subject Sample # 2	0.211	02:16
Breath Volume	1.953	02:16
Air Blank	0.000	02:17
Control Test	0.079	02:17
Air Blank	0.000	02:18
Diagnostics Check	OK	02:18

Data Download Date/Time: 06/27/2017 14:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIGHTSEY,ROBERT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: STANLEY,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 06/17/2017 03:31:58

Diagnostics Check	OK	03:25
Air Blank	0.000	03:26
Control Test	0.080	03:26
Air Blank	0.000	03:27
Subject Sample # 1	0.121	03:27
Breath Volume	1.820	03:27
Air Blank	0.000	03:28
Air Blank	0.000	03:30
Subject Sample # 2	0.121	03:30
Breath Volume	2.406	03:30
Air Blank	0.000	03:31
Control Test	0.080	03:31
Air Blank	0.000	03:31
Diagnostics Check	OK	03:31

Data Download Date/Time: 06/27/2017 14:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIGHTSEY,ROBERT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CAMACHO,K
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 06/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:03

SUBJECT'S LAST NAME: RIVERA MEJIAS

DATE/TIME OF TEST: 06/17/2017 04:30:24

Diagnostics Check	OK	04:23
Air Blank	0.000	04:24
Control Test	0.080	04:24
Air Blank	0.000	04:25
Subject Sample # 1	0.129	04:25
Breath Volume	2.933	04:25
Air Blank	0.000	04:26
Air Blank	0.000	04:28
Subject Sample # 2	0.128	04:28
Breath Volume	2.128	04:28
Air Blank	0.000	04:29
Control Test	0.080	04:29
Air Blank	0.000	04:30
Diagnostics Check	OK	04:30

Data Download Date/Time: 06/27/2017 14:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIGHTSEY,ROBERT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SWARTZ,W
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 06/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:32

SUBJECT'S LAST NAME: KISER

DATE/TIME OF TEST: 06/17/2017 05:58:50

Diagnostics Check	OK	05:52
Air Blank	0.000	05:52
Control Test	0.080	05:53
Air Blank	0.000	05:53
Subject Sample # 1	0.125	05:54
Breath Volume	1.792	05:54
Air Blank	0.000	05:54
Air Blank	0.000	05:56
Subject Sample # 2	0.122	05:57
Breath Volume	1.601	05:57
Air Blank	0.000	05:57
Control Test	0.080	05:58
Air Blank	0.000	05:58
Diagnostics Check	OK	05:58

Data Download Date/Time: 06/27/2017 14:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS, CHRISTOPHER, D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: RODRIGUEZ, M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:27

SUBJECT'S LAST NAME: VOSS

DATE/TIME OF TEST: 06/18/2017 01:59:08

Diagnostics Check	OK	01:47	
Air Blank	0.000	01:48	
Control Test	0.080	01:48	
Air Blank	0.000	01:48	
Subject Sample	# 1 0.089	01:51	Volume Not Met
Breath Volume	0.820	01:51	
Air Blank	0.000	01:52	
Air Blank	0.000	01:54	
Subject Sample	# 2 0.088	01:57	Volume Not Met
Breath Volume	0.933	01:57	
Air Blank	0.000	01:58	
Control Test	0.078	01:58	
Air Blank	0.000	01:59	
Diagnostics Check	OK	01:59	

Data Download Date/Time: 06/27/2017 14:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS, CHRISTOPHER, D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: STANLEY, T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:33

SUBJECT'S LAST NAME: HIRABAYASHI

DATE/TIME OF TEST: 06/18/2017 03:02:40

Diagnostics Check	OK	02:56
Air Blank	0.000	02:56
Control Test	0.080	02:57
Air Blank	0.000	02:57
Subject Sample # 1	0.093	02:58
Breath Volume	3.492	02:58
Air Blank	0.000	02:58
Air Blank	0.000	03:00
Subject Sample # 2	0.094	03:01
Breath Volume	3.714	03:01
Air Blank	0.000	03:01
Control Test	0.080	03:02
Air Blank	0.000	03:02
Diagnostics Check	OK	03:02

Data Download Date/Time: 06/27/2017 14:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CHERENA,F
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:55

SUBJECT'S LAST NAME: MCPHERSON

DATE/TIME OF TEST: 06/18/2017 05:18:37

Diagnostics Check	OK	05:15	
Air Blank	0.000	05:16	
Control Test	0.080	05:16	
Air Blank	0.000	05:17	
Subject Sample	# 1	05:17	Subject Test Refused
Breath Volume	0.000	05:17	
Air Blank	0.000	05:17	
Control Test	0.079	05:18	
Air Blank	0.000	05:18	
Diagnostics Check	OK	05:18	

Data Download Date/Time: 06/27/2017 14:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MORRISON,M
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 06/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:42

SUBJECT'S LAST NAME: COLLERD

DATE/TIME OF TEST: 06/18/2017 06:07:54

Diagnostics Check	OK	06:03	
Air Blank	0.000	06:04	
Control Test	0.079	06:04	
Air Blank	0.000	06:04	
Subject Sample	# 1	06:06	Subject Test Refused
Breath Volume	0.000	06:06	
Air Blank	0.000	06:06	
Control Test	0.080	06:07	
Air Blank	0.000	06:07	
Diagnostics Check	OK	06:07	

Data Download Date/Time: 06/27/2017 14:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FARLEY,J
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 06/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:19

SUBJECT'S LAST NAME: SUTTON

DATE/TIME OF TEST: 06/19/2017 03:43:01

Diagnostics Check	OK	03:40	
Air Blank	0.000	03:40	
Control Test	0.080	03:40	
Air Blank	0.000	03:41	
Subject Sample	# 1	03:41	Subject Test Refused
Breath Volume	0.000	03:41	
Air Blank	0.000	03:42	
Control Test	0.081	03:42	
Air Blank	0.000	03:42	
Diagnostics Check	OK	03:43	

Data Download Date/Time: 06/27/2017 14:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIGHTSEY,ROBERT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CHRISTOFFERSEN,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:07

SUBJECT'S LAST NAME: ALCANTARA

DATE/TIME OF TEST: 06/19/2017 04:37:58

Diagnostics Check	OK	04:27
Air Blank	0.000	04:28
Control Test	0.079	04:28
Air Blank	0.000	04:29
Subject Sample # 1	0.227	04:29
Breath Volume	2.023	04:29
Air Blank	0.000	04:30
Air Blank	0.000	04:32
Subject Sample # 2	0.204	04:33
Breath Volume	1.464	04:33
Air Blank	0.000	04:33
Air Blank	0.000	04:35
Subject Sample # 3	0.203	04:36
Breath Volume	1.578	04:36
Air Blank	0.000	04:37
Control Test	0.078	04:37
Air Blank	0.000	04:37
Diagnostics Check	OK	04:37

Data Download Date/Time: 06/27/2017 14:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: COHILL,D
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 06/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:49

SUBJECT'S LAST NAME: LINK

DATE/TIME OF TEST: 06/20/2017 02:15:58

Diagnostics Check	OK	02:09
Air Blank	0.000	02:10
Control Test	0.080	02:10
Air Blank	0.000	02:10
Subject Sample # 1	0.141	02:11
Breath Volume	2.710	02:11
Air Blank	0.000	02:12
Air Blank	0.000	02:13
Subject Sample # 2	0.141	02:14
Breath Volume	2.640	02:14
Air Blank	0.000	02:15
Control Test	0.079	02:15
Air Blank	0.000	02:15
Diagnostics Check	OK	02:15

Data Download Date/Time: 06/27/2017 14:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BRETT,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:36

SUBJECT'S LAST NAME: CANO

DATE/TIME OF TEST: 06/20/2017 05:03:08

Diagnostics Check	OK	04:56
Air Blank	0.000	04:57
Control Test	0.080	04:57
Air Blank	0.000	04:58
Subject Sample # 1	0.122	04:58
Breath Volume	1.457	04:58
Air Blank	0.000	04:59
Air Blank	0.000	05:01
Subject Sample # 2	0.122	05:01
Breath Volume	2.742	05:01
Air Blank	0.000	05:02
Control Test	0.079	05:02
Air Blank	0.000	05:03
Diagnostics Check	OK	05:03

Data Download Date/Time: 06/27/2017 14:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MULLIS,M,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BRYANT,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:13

SUBJECT'S LAST NAME: CHANEY

DATE/TIME OF TEST: 06/22/2017 01:37:03

Diagnostics Check	OK	01:34	
Air Blank	0.000	01:34	
Control Test	0.080	01:34	
Air Blank	0.000	01:35	
Subject Sample	# 1	01:35	Subject Test Refused
Breath Volume	0.000	01:35	
Air Blank	0.000	01:36	
Control Test	0.080	01:36	
Air Blank	0.000	01:36	
Diagnostics Check	OK	01:37	

Data Download Date/Time: 06/27/2017 14:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBER,JOSEPH,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BYRNE,J.F.
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:14

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 06/22/2017 02:41:04

Diagnostics Check	OK	02:34
Air Blank	0.000	02:35
Control Test	0.079	02:35
Air Blank	0.000	02:36
Subject Sample # 1	0.215	02:36
Breath Volume	2.683	02:36
Air Blank	0.000	02:37
Air Blank	0.000	02:39
Subject Sample # 2	0.215	02:39
Breath Volume	2.718	02:39
Air Blank	0.000	02:40
Control Test	0.079	02:40
Air Blank	0.000	02:40
Diagnostics Check	OK	02:41

Data Download Date/Time: 06/27/2017 14:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBER,JOSEPH,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FILBERT,E.D.
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 06/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: DOBSON

DATE/TIME OF TEST: 06/22/2017 04:57:21

Diagnostics Check	OK	04:50
Air Blank	0.000	04:50
Control Test	0.080	04:51
Air Blank	0.000	04:51
Subject Sample # 1	0.096	04:52
Breath Volume	1.773	04:52
Air Blank	0.000	04:52
Air Blank	0.000	04:54
Subject Sample # 2	0.094	04:55
Breath Volume	1.792	04:55
Air Blank	0.000	04:56
Control Test	0.080	04:56
Air Blank	0.000	04:57
Diagnostics Check	OK	04:57

Data Download Date/Time: 06/27/2017 14:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON, BRADEN, J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: TORRES, W
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 06/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:17

SUBJECT'S LAST NAME: ADAMS

DATE/TIME OF TEST: 06/23/2017 02:45:47

Diagnostics Check	OK	02:39
Air Blank	0.000	02:39
Control Test	0.079	02:39
Air Blank	0.000	02:40
Subject Sample # 1	0.106	02:41
Breath Volume	1.601	02:41
Air Blank	0.000	02:41
Air Blank	0.000	02:43
Subject Sample # 2	0.096	02:44
Breath Volume	1.714	02:44
Air Blank	0.000	02:44
Control Test	0.079	02:45
Air Blank	0.000	02:45
Diagnostics Check	OK	02:45

Data Download Date/Time: 06/27/2017 14:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON, BRADEN, J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: THOMAS, G
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:33

SUBJECT'S LAST NAME: GEIGER

DATE/TIME OF TEST: 06/23/2017 05:00:09

Diagnostics Check	OK	04:53
Air Blank	0.000	04:54
Control Test	0.080	04:54
Air Blank	0.000	04:55
Subject Sample # 1	0.161	04:55
Breath Volume	3.105	04:55
Air Blank	0.000	04:56
Air Blank	0.000	04:57
Subject Sample # 2	0.164	04:58
Breath Volume	3.066	04:58
Air Blank	0.000	04:59
Control Test	0.079	04:59
Air Blank	0.000	05:00
Diagnostics Check	OK	05:00

Data Download Date/Time: 06/27/2017 14:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON, BRADEN, J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: TORRES, W
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 06/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 05:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:55

SUBJECT'S LAST NAME: SAMARDZIC

DATE/TIME OF TEST: 06/23/2017 06:18:13

Diagnostics Check	OK	06:15	
Air Blank	0.000	06:15	
Control Test	0.080	06:16	
Air Blank	0.000	06:16	
Subject Sample	# 1	06:16	Subject Test Refused
Breath Volume	0.000	06:16	
Air Blank	0.000	06:17	
Control Test	0.080	06:17	
Air Blank	0.000	06:18	
Diagnostics Check	OK	06:18	

Data Download Date/Time: 06/27/2017 14:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLARIDA,K,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: RUIZ,R
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 06/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 12:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:15

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 06/23/2017 14:42:08

Diagnostics Check	OK	14:36
Air Blank	0.000	14:36
Control Test	0.080	14:36
Air Blank	0.000	14:37
Subject Sample # 1	0.011	14:37
Breath Volume	3.363	14:37
Air Blank	0.000	14:38
Air Blank	0.000	14:40
Subject Sample # 2	0.011	14:40
Breath Volume	3.320	14:40
Air Blank	0.000	14:41
Control Test	0.081	14:41
Air Blank	0.000	14:42
Diagnostics Check	OK	14:42

Data Download Date/Time: 06/27/2017 14:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTELL,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: GARCIA,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 17:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:15

SUBJECT'S LAST NAME: WILKS

DATE/TIME OF TEST: 06/24/2017 18:47:23

Diagnostics Check	OK	18:36
Air Blank	0.000	18:36
Control Test	0.079	18:36
Air Blank	0.000	18:37
Subject Sample # 1	0.235	18:38
Breath Volume	1.175	18:38
Air Blank	0.000	18:39
Air Blank	0.000	18:41
Subject Sample # 2	0.198	18:42
Breath Volume	1.304	18:42
Air Blank	0.000	18:43
Air Blank	0.000	18:45
Subject Sample # 3	0.239	18:45
Breath Volume	2.011	18:45
Air Blank	0.000	18:46
Control Test	0.077	18:46
Air Blank	0.000	18:47
Diagnostics Check	OK	18:47

Data Download Date/Time: 06/27/2017 14:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTELL,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SHEARER,T
ARRESTING OFFICER AGENCY: FFWCC (Lake City)
LAST AGENCY INSPECTION: 06/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:25

SUBJECT'S LAST NAME: BECKFIELD

DATE/TIME OF TEST: 06/24/2017 22:48:37

Diagnostics Check	OK	22:45	
Air Blank	0.000	22:46	
Control Test	0.079	22:46	
Air Blank	0.000	22:46	
Subject Sample	# 1	22:47	Subject Test Refused
Breath Volume	0.000	22:47	
Air Blank	0.000	22:47	
Control Test	0.080	22:48	
Air Blank	0.000	22:48	
Diagnostics Check	OK	22:48	

Data Download Date/Time: 06/27/2017 14:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURNS,D,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MORRISON,M
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 06/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:03

SUBJECT'S LAST NAME: PADGHAM

DATE/TIME OF TEST: 06/27/2017 04:27:19

Diagnostics Check	OK	04:24	
Air Blank	0.000	04:24	
Control Test	0.079	04:25	
Air Blank	0.000	04:25	
Subject Sample	# 1	04:25	Subject Test Refused
Breath Volume	0.000	04:25	
Air Blank	0.000	04:26	
Control Test	0.080	04:26	
Air Blank	0.000	04:27	
Diagnostics Check	OK	04:27	

Data Download Date/Time: 06/27/2017 14:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBER,JOSEPH,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MOSLEY,M.R.
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:13

SUBJECT'S LAST NAME: GJOKA

DATE/TIME OF TEST: 06/28/2017 00:36:51

Diagnostics Check	OK	00:34	
Air Blank	0.000	00:34	
Control Test	0.080	00:34	
Air Blank	0.000	00:35	
Subject Sample	# 1	00:35	Subject Test Refused
Breath Volume	0.000	00:35	
Air Blank	0.000	00:35	
Control Test	0.080	00:36	
Air Blank	0.000	00:36	
Diagnostics Check	OK	00:36	

Data Download Date/Time: 07/13/2017 15:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBER,JOSEPH,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: GARCIA,J.R.
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:31

SUBJECT'S LAST NAME: CHURCH

DATE/TIME OF TEST: 06/28/2017 03:56:52

Diagnostics Check	OK	03:53	
Air Blank	0.000	03:54	
Control Test	0.080	03:54	
Air Blank	0.000	03:55	
Subject Sample	# 1	03:55	Subject Test Refused
Breath Volume	0.000	03:55	
Air Blank	0.000	03:55	
Control Test	0.079	03:56	
Air Blank	0.000	03:56	
Diagnostics Check	OK	03:56	

Data Download Date/Time: 07/13/2017 15:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBER,JOSEPH,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: STANLEY,T.L.
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:19

SUBJECT'S LAST NAME: WAGNER

DATE/TIME OF TEST: 06/29/2017 00:46:30

Diagnostics Check	OK	00:39
Air Blank	0.000	00:40
Control Test	0.080	00:40
Air Blank	0.000	00:41
Subject Sample # 1	0.055	00:42
Breath Volume	3.031	00:42
Air Blank	0.000	00:42
Air Blank	0.000	00:44
Subject Sample # 2	0.056	00:45
Breath Volume	2.898	00:45
Air Blank	0.000	00:45
Control Test	0.080	00:45
Air Blank	0.000	00:46
Diagnostics Check	OK	00:46

Data Download Date/Time: 07/13/2017 15:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBER,JOSEPH,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: EANES,J.L.
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:13

SUBJECT'S LAST NAME: PRATT

DATE/TIME OF TEST: 06/30/2017 01:40:21

Diagnostics Check	OK	01:33
Air Blank	0.000	01:33
Control Test	0.080	01:34
Air Blank	0.000	01:34
Subject Sample # 1	0.077	01:35
Breath Volume	1.457	01:35
Air Blank	0.000	01:36
Air Blank	0.000	01:38
Subject Sample # 2	0.085	01:38
Breath Volume	2.011	01:38
Air Blank	0.000	01:39
Control Test	0.080	01:39
Air Blank	0.000	01:40
Diagnostics Check	OK	01:40

Data Download Date/Time: 07/13/2017 15:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBER,JOSEPH,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BROWN,T.A.
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:42

SUBJECT'S LAST NAME: LEE

DATE/TIME OF TEST: 06/30/2017 03:15:00

Diagnostics Check	OK	03:02	
Air Blank	0.000	03:03	
Control Test	0.079	03:03	
Air Blank	0.000	03:03	
Subject Sample	# 1 0.234	03:04	
Breath Volume	2.085	03:04	
Air Blank	0.000	03:05	
Air Blank	0.000	03:07	
Subject Sample	# 2 0.212	03:10	Slope Not Level
Breath Volume	1.832	03:10	
Air Blank	0.000	03:11	
Air Blank	0.000	03:12	
Subject Sample	# 3 0.222	03:13	
Breath Volume	1.960	03:13	
Air Blank	0.000	03:14	
Control Test	0.077	03:14	
Air Blank	0.000	03:14	
Diagnostics Check	OK	03:15	

Data Download Date/Time: 07/13/2017 15:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBER,JOSEPH,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: TORRES,W.E.
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 06/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:09

SUBJECT'S LAST NAME: ORTIZ

DATE/TIME OF TEST: 06/30/2017 23:36:16

Diagnostics Check	OK	23:29
Air Blank	0.000	23:30
Control Test	0.080	23:30
Air Blank	0.000	23:31
Subject Sample # 1	0.134	23:31
Breath Volume	2.601	23:31
Air Blank	0.000	23:32
Air Blank	0.000	23:34
Subject Sample # 2	0.135	23:34
Breath Volume	2.691	23:34
Air Blank	0.000	23:35
Control Test	0.079	23:35
Air Blank	0.000	23:36
Diagnostics Check	OK	23:36

Data Download Date/Time: 07/13/2017 15:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001865

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUACKENBUSH,BRET,C
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: QUACKENBUSH,BRET
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 05/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:36

SUBJECT'S LAST NAME: GORDON

DATE/TIME OF TEST: 06/04/2017 02:04:35

Diagnostics Check	OK	01:58
Air Blank	0.000	01:58
Control Test	0.075	01:58
Air Blank	0.000	01:59
Subject Sample # 1	0.000	02:00
Breath Volume	1.304	02:00
Air Blank	0.000	02:00
Air Blank	0.000	02:02
Subject Sample # 2	0.000	02:03
Breath Volume	2.109	02:03
Air Blank	0.000	02:03
Control Test	0.075	02:04
Air Blank	0.000	02:04
Diagnostics Check	OK	02:04

Data Download Date/Time: 06/21/2017 10:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001865

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KELLNER, MICHAEL, A
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: KELLNER, MICHAEL
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 05/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:15

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 06/07/2017 20:59:32

Diagnostics Check	OK	20:48	
Air Blank	0.000	20:48	
Control Test	0.077	20:48	
Air Blank	0.000	20:49	
Subject Sample	# 1	20:52	No Sample Provided
Breath Volume	0.000	20:52	
Air Blank	0.000	20:53	
Air Blank	0.000	20:55	
Subject Sample	# 2	20:58	No Sample Provided
Breath Volume	0.000	20:58	
Air Blank	0.000	20:58	
Control Test	0.078	20:58	
Air Blank	0.000	20:59	
Diagnostics Check	OK	20:59	

Data Download Date/Time: 06/21/2017 10:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001865

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUACKENBUSH,BRET,C
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: QUACKENBUSH,BRET
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 05/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:29

SUBJECT'S LAST NAME: DRIVER

DATE/TIME OF TEST: 06/08/2017 03:56:33

Diagnostics Check	OK	03:49
Air Blank	0.000	03:49
Control Test	0.077	03:50
Air Blank	0.000	03:50
Subject Sample # 1	0.079	03:51
Breath Volume	2.070	03:51
Air Blank	0.000	03:52
Air Blank	0.000	03:54
Subject Sample # 2	0.078	03:55
Breath Volume	1.542	03:55
Air Blank	0.000	03:55
Control Test	0.077	03:55
Air Blank	0.000	03:56
Diagnostics Check	OK	03:56

Data Download Date/Time: 06/21/2017 10:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001865

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KELLNER, MICHAEL, A
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: KELLNER, MICHAEL
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 05/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:04

SUBJECT'S LAST NAME: DUNLEVY

DATE/TIME OF TEST: 06/09/2017 04:55:37

Diagnostics Check	OK	04:45	
Air Blank	0.000	04:46	
Control Test	0.077	04:46	
Air Blank	0.000	04:47	
Subject Sample	# 1	04:50	No Sample Provided
Breath Volume	0.000	04:50	
Air Blank	0.000	04:50	
Air Blank	0.000	04:52	
Subject Sample	# 2	04:54	RFI Detect
Breath Volume	0.000	04:54	
Air Blank	0.000	04:54	
Control Test	0.078	04:55	
Air Blank	0.000	04:55	
Diagnostics Check	OK	04:55	

Data Download Date/Time: 06/21/2017 10:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001865

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KELLNER, MICHAEL, A
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: KELLNER, MICHAEL
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 05/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:04

SUBJECT'S LAST NAME: DUNLEVY

DATE/TIME OF TEST: 06/09/2017 05:11:46

Diagnostics Check	OK	05:00	
Air Blank	0.000	05:00	
Control Test	0.077	05:01	
Air Blank	0.000	05:01	
Subject Sample	# 1	05:04	No Sample Provided
Breath Volume	0.000	05:04	
Air Blank	0.000	05:05	
Air Blank	0.000	05:07	
Subject Sample	# 2	05:10	No Sample Provided
Breath Volume	0.000	05:10	
Air Blank	0.000	05:10	
Control Test	0.077	05:11	
Air Blank	0.000	05:11	
Diagnostics Check	OK	05:11	

Data Download Date/Time: 06/21/2017 10:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001865

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTLE,CORY,A
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: SUTTLE,CORY
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 05/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: PICKETT

DATE/TIME OF TEST: 06/11/2017 01:26:52

Diagnostics Check	OK	01:20
Air Blank	0.000	01:21
Control Test	0.078	01:21
Air Blank	0.000	01:22
Subject Sample # 1	0.129	01:22
Breath Volume	1.578	01:22
Air Blank	0.000	01:23
Air Blank	0.000	01:25
Subject Sample # 2	0.134	01:25
Breath Volume	1.730	01:25
Air Blank	0.000	01:25
Control Test	0.077	01:26
Air Blank	0.000	01:26
Diagnostics Check	OK	01:26

Data Download Date/Time: 06/21/2017 10:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001865

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTLE,CORY,A
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: SUTTLE,CORY
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 05/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: GOFFE

DATE/TIME OF TEST: 06/11/2017 01:59:59

Diagnostics Check	OK	01:51	
Air Blank	0.000	01:51	
Control Test	0.078	01:51	
Air Blank	0.000	01:52	
Subject Sample	# 1	01:55	No Sample Provided
Breath Volume	0.000	01:55	
Air Blank	0.000	01:55	
Air Blank	0.000	01:57	
Subject Sample	# 2	01:58	Subject Test Refused
Breath Volume	0.000	01:58	
Air Blank	0.000	01:59	
Control Test	0.078	01:59	
Air Blank	0.000	01:59	
Diagnostics Check	OK	01:59	

Data Download Date/Time: 06/21/2017 10:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001865

INSTRUMENT REGISTERED TO: LAKELAND PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SUTTLE,CORY,A

BREATH TEST OPERATOR AGENCY: Lakeland PD

ARRESTING OFFICER: SUTTLE,CORY

ARRESTING OFFICER AGENCY: Lakeland PD

LAST AGENCY INSPECTION: 05/17/2017

Date or Time Changed: No

DRY GAS STANDARD LOT #: 22315080A2

DRY GAS EXPIRY DATE: 09/05/2017

VIOLATION CODE: DUI

ARREST TIME: 00:45

GENDER: Female

OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: GOFFE

DATE/TIME OF TEST: 06/11/2017 02:09:32

Diagnostics Check	OK	02:06	
Air Blank	0.000	02:06	
Control Test	0.078	02:07	
Air Blank	0.000	02:07	
Subject Sample	# 1	02:08	Subject Test Refused
Breath Volume	0.000	02:08	
Air Blank	0.000	02:08	
Control Test	0.077	02:08	
Air Blank	0.000	02:09	
Diagnostics Check	OK	02:09	

Data Download Date/Time: 06/21/2017 10:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001865

INSTRUMENT REGISTERED TO: LAKELAND PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MULDERRIG,SEAN,F

BREATH TEST OPERATOR AGENCY: Lakeland PD

ARRESTING OFFICER: MULDERRIG,SEAN

ARRESTING OFFICER AGENCY: Lakeland PD

LAST AGENCY INSPECTION: 05/17/2017

Date or Time Changed: No

DRY GAS STANDARD LOT #: 22315080A2

DRY GAS EXPIRY DATE: 09/05/2017

VIOLATION CODE: DUI

ARREST TIME: 04:19

GENDER: Male

OBSERVATION PERIOD BEGAN: 04:35

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 06/11/2017 05:06:36

Diagnostics Check	OK	04:55	
Air Blank	0.000	04:55	
Control Test	0.077	04:56	
Air Blank	0.000	04:56	
Subject Sample	# 1	04:59	No Sample Provided
Breath Volume	0.000	04:59	
Air Blank	0.000	05:00	
Air Blank	0.000	05:02	
Subject Sample	# 2	05:05	No Sample Provided
Breath Volume	0.000	05:05	
Air Blank	0.000	05:05	
Control Test	0.079	05:06	
Air Blank	0.000	05:06	
Diagnostics Check	OK	05:06	

Data Download Date/Time: 06/21/2017 10:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001865

INSTRUMENT REGISTERED TO: LAKELAND PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MULDERRIG,SEAN,F

BREATH TEST OPERATOR AGENCY: Lakeland PD

ARRESTING OFFICER: MULDERRIG,SEAN

ARRESTING OFFICER AGENCY: Lakeland PD

LAST AGENCY INSPECTION: 05/17/2017

Date or Time Changed: No

DRY GAS STANDARD LOT #: 22315080A2

DRY GAS EXPIRY DATE: 09/05/2017

VIOLATION CODE: DUI

ARREST TIME: 02:39

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:26

SUBJECT'S LAST NAME: CHAPMAN

DATE/TIME OF TEST: 06/15/2017 03:59:13

Diagnostics Check	OK	03:47	
Air Blank	0.000	03:48	
Control Test	0.077	03:48	
Air Blank	0.000	03:49	
Subject Sample	# 1	03:52	No Sample Provided
Breath Volume	0.000	03:52	
Air Blank	0.000	03:52	
Air Blank	0.000	03:54	
Subject Sample	# 2	03:57	No Sample Provided
Breath Volume	0.000	03:57	
Air Blank	0.000	03:58	
Control Test	0.078	03:58	
Air Blank	0.000	03:59	
Diagnostics Check	OK	03:59	

Data Download Date/Time: 06/21/2017 10:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001865

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SULLIVAN,MELANIE,A
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: SULLIVAN,MELANIE
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 05/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:33

SUBJECT'S LAST NAME: KINNETT

DATE/TIME OF TEST: 06/17/2017 04:06:23

Diagnostics Check	OK	03:59
Air Blank	0.000	04:00
Control Test	0.077	04:00
Air Blank	0.000	04:01
Subject Sample # 1	0.171	04:01
Breath Volume	2.390	04:01
Air Blank	0.000	04:02
Air Blank	0.000	04:04
Subject Sample # 2	0.168	04:04
Breath Volume	2.218	04:04
Air Blank	0.000	04:05
Control Test	0.077	04:05
Air Blank	0.000	04:06
Diagnostics Check	OK	04:06

Data Download Date/Time: 06/21/2017 10:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001865

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KELLNER, MICHAEL, A
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: KELLNER, MICHAEL
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 05/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: LINSENMANN

DATE/TIME OF TEST: 06/19/2017 02:11:56

Diagnostics Check	OK	02:04
Air Blank	0.000	02:04
Control Test	0.078	02:05
Air Blank	0.000	02:05
Subject Sample # 1	0.199	02:06
Breath Volume	1.441	02:06
Air Blank	0.000	02:07
Air Blank	0.000	02:08
Subject Sample # 2	0.196	02:10
Breath Volume	2.015	02:10
Air Blank	0.000	02:11
Control Test	0.078	02:11
Air Blank	0.000	02:11
Diagnostics Check	OK	02:11

Data Download Date/Time: 06/21/2017 10:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002097

INSTRUMENT REGISTERED TO: NAS JACKSONVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARKER,TIMOTHY,N
BREATH TEST OPERATOR AGENCY: NAS Jacksonville PD
ARRESTING OFFICER: PARKER,TIMOTHY
ARRESTING OFFICER AGENCY: NAS Jacksonville PD
LAST AGENCY INSPECTION: 06/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: ADMIN
ARREST TIME: 15:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:40

SUBJECT'S LAST NAME: POTTS

DATE/TIME OF TEST: 06/21/2017 17:35:36

Diagnostics Check	OK	17:29
Air Blank	0.000	17:30
Control Test	0.079	17:30
Air Blank	0.000	17:30
Subject Sample # 1	0.010	17:31
Breath Volume	2.464	17:31
Air Blank	0.000	17:31
Air Blank	0.000	17:33
Subject Sample # 2	0.010	17:34
Breath Volume	2.382	17:34
Air Blank	0.000	17:34
Control Test	0.079	17:34
Air Blank	0.000	17:35
Diagnostics Check	OK	17:35

Data Download Date/Time: 07/20/2017 02:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002170

INSTRUMENT REGISTERED TO: NAS KEY WEST
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANDERSON,LEE,B
BREATH TEST OPERATOR AGENCY: NAS Key West
ARRESTING OFFICER: ANDERSON,LEE
ARRESTING OFFICER AGENCY: NAS Key West
LAST AGENCY INSPECTION: 06/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1514466
DRY GAS EXPIRY DATE: 06/18/2016
VIOLATION CODE: ZERO TOL
ARREST TIME: 12:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:35

SUBJECT'S LAST NAME: HARDAMAN

DATE/TIME OF TEST: 06/15/2017 13:02:10

Diagnostics Check	OK	12:55
Air Blank	0.000	12:56
Control Test	0.080	12:56
Air Blank	0.000	12:56
Subject Sample # 1	0.065	12:57
Breath Volume	3.816	12:57
Air Blank	0.000	12:58
Air Blank	0.000	13:00
Subject Sample # 2	0.063	13:00
Breath Volume	4.265	13:00
Air Blank	0.000	13:01
Control Test	0.079	13:01
Air Blank	0.000	13:02
Diagnostics Check	OK	13:02

Data Download Date/Time: 08/01/2017 11:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002170

INSTRUMENT REGISTERED TO: NAS KEY WEST
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANDERSON,LEE,B
BREATH TEST OPERATOR AGENCY: NAS Key West
ARRESTING OFFICER: ANDERSON,LEE
ARRESTING OFFICER AGENCY: NAS Key West
LAST AGENCY INSPECTION: 06/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1514466
DRY GAS EXPIRY DATE: 06/18/2016
VIOLATION CODE: ZERO TOL
ARREST TIME: 12:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:35

SUBJECT'S LAST NAME: AMERSON

DATE/TIME OF TEST: 06/15/2017 13:16:04

Diagnostics Check	OK	13:09
Air Blank	0.000	13:10
Control Test	0.080	13:10
Air Blank	0.000	13:10
Subject Sample # 1	0.063	13:11
Breath Volume	3.656	13:11
Air Blank	0.000	13:12
Air Blank	0.000	13:14
Subject Sample # 2	0.059	13:14
Breath Volume	3.503	13:14
Air Blank	0.000	13:15
Control Test	0.079	13:15
Air Blank	0.000	13:15
Diagnostics Check	OK	13:16

Data Download Date/Time: 08/01/2017 11:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002231

INSTRUMENT REGISTERED TO: NEW SMYRNA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,DAVID,B
BREATH TEST OPERATOR AGENCY: New Smyrna Beach PD
ARRESTING OFFICER: POTTER,KYLE
ARRESTING OFFICER AGENCY: Edgewater PD
LAST AGENCY INSPECTION: 05/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A3
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 10:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:38

SUBJECT'S LAST NAME: KNOESS

DATE/TIME OF TEST: 06/10/2017 23:01:26

Diagnostics Check	OK	22:58	
Air Blank	0.000	22:58	
Control Test	0.080	22:59	
Air Blank	0.000	22:59	
Subject Sample	# 1	23:00	Subject Test Refused
Breath Volume	0.000	23:00	
Air Blank	0.000	23:00	
Control Test	0.080	23:00	
Air Blank	0.000	23:01	
Diagnostics Check	OK	23:01	

Data Download Date/Time: 06/28/2017 16:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002231

INSTRUMENT REGISTERED TO: NEW SMYRNA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,DAVID,B
BREATH TEST OPERATOR AGENCY: New Smyrna Beach PD
ARRESTING OFFICER: JENKINS,ANDRE
ARRESTING OFFICER AGENCY: New Smyrna Beach PD
LAST AGENCY INSPECTION: 05/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A3
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: LAEL

DATE/TIME OF TEST: 06/15/2017 02:13:13

Diagnostics Check	OK	02:05
Air Blank	0.000	02:05
Control Test	0.080	02:06
Air Blank	0.000	02:06
Subject Sample # 1	0.000	02:07
Breath Volume	1.820	02:07
Air Blank	0.000	02:08
Air Blank	0.000	02:10
Subject Sample # 2	0.000	02:11
Breath Volume	1.566	02:11
Air Blank	0.000	02:12
Control Test	0.079	02:12
Air Blank	0.000	02:13
Diagnostics Check	OK	02:13

Data Download Date/Time: 06/28/2017 16:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002462

INSTRUMENT REGISTERED TO: HIALEAH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRAND,ANDREW,
BREATH TEST OPERATOR AGENCY: Hialeah PD
ARRESTING OFFICER: RODRIGUEZ,YAIDEL
ARRESTING OFFICER AGENCY: Hialeah PD
LAST AGENCY INSPECTION: 05/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 712031
DRY GAS EXPIRY DATE: 02/21/2020
VIOLATION CODE: DUI
ARREST TIME: 22:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:54

SUBJECT'S LAST NAME: MARCHAN

DATE/TIME OF TEST: 06/09/2017 23:17:10

Diagnostics Check	OK	23:14	
Air Blank	0.000	23:14	
Control Test	0.080	23:15	
Air Blank	0.000	23:15	
Subject Sample	# 1	23:15	Subject Test Refused
Breath Volume	0.000	23:15	
Air Blank	0.000	23:16	
Control Test	0.079	23:16	
Air Blank	0.000	23:17	
Diagnostics Check	OK	23:17	

Data Download Date/Time: 06/22/2017 12:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002989

INSTRUMENT REGISTERED TO: TALLAHASSEE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROY,G,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: GARRETT,S
ARRESTING OFFICER AGENCY: Leon County SO
LAST AGENCY INSPECTION: 05/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:53

SUBJECT'S LAST NAME: AUSTIN

DATE/TIME OF TEST: 06/05/2017 22:34:29

Diagnostics Check	OK	22:26
Air Blank	0.000	22:26
Control Test	0.078	22:27
Air Blank	0.000	22:27
Subject Sample # 1	0.204	22:29
Breath Volume	1.613	22:29
Air Blank	0.000	22:30
Air Blank	0.000	22:31
Subject Sample # 2	0.207	22:32
Breath Volume	2.304	22:32
Air Blank	0.000	22:33
Control Test	0.078	22:33
Air Blank	0.000	22:34
Diagnostics Check	OK	22:34

Data Download Date/Time: 06/30/2017 21:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002989

INSTRUMENT REGISTERED TO: TALLAHASSEE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KELLER, DAVID,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: KELLER, DAVID
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 05/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: QASSIS

DATE/TIME OF TEST: 06/16/2017 02:08:41

Diagnostics Check	OK	01:56	
Air Blank	0.000	01:57	
Control Test	0.080	01:57	
Air Blank	0.000	01:58	
Subject Sample	# 1 0.156	02:01	Volume Not Met
Breath Volume	0.187	02:01	
Air Blank	0.000	02:02	
Air Blank	0.000	02:03	
Subject Sample	# 2 0.136	02:07	Volume Not Met
Breath Volume	0.250	02:07	
Air Blank	0.000	02:07	
Control Test	0.078	02:08	
Air Blank	0.000	02:08	
Diagnostics Check	OK	02:08	

Data Download Date/Time: 06/30/2017 21:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003195

INSTRUMENT REGISTERED TO: FWC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KEARNEY,BRANDON,G
BREATH TEST OPERATOR AGENCY: FFWCC (Ocala)
ARRESTING OFFICER: KEARNEY,BRANDON
ARRESTING OFFICER AGENCY: FFWCC (Ocala)
LAST AGENCY INSPECTION: 06/03/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: BUI
ARREST TIME: 19:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: DOE

DATE/TIME OF TEST: 06/10/2017 20:31:25

Diagnostics Check	OK	20:25
Air Blank	0.000	20:25
Control Test	0.079	20:26
Air Blank	0.000	20:26
Subject Sample # 1	0.000	20:26
Breath Volume	2.351	20:26
Air Blank	0.000	20:27
Air Blank	0.000	20:29
Subject Sample # 2	0.000	20:30
Breath Volume	1.472	20:30
Air Blank	0.000	20:30
Control Test	0.079	20:30
Air Blank	0.000	20:31
Diagnostics Check	OK	20:31

Data Download Date/Time: 08/02/2017 05:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003197

INSTRUMENT REGISTERED TO: FWC	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 477627
BREATH TEST OPERATOR: MOLNAR,JARROD,T	DRY GAS EXPIRY DATE: 06/29/2019
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)	VIOLATION CODE: BUI
ARRESTING OFFICER: TISON,TAYLOR	ARREST TIME: 14:19
ARRESTING OFFICER AGENCY: FFWCC (Panama City)	GENDER: Male
LAST AGENCY INSPECTION: 05/11/2017	OBSERVATION PERIOD BEGAN: 14:19

SUBJECT'S LAST NAME: HOLDER DATE/TIME OF TEST: 06/03/2017 14:51:18

Diagnostics Check	OK	14:48	
Air Blank	0.000	14:48	
Control Test	0.080	14:49	
Air Blank	0.000	14:49	
Subject Sample	# 1	14:49	Subject Test Refused
Breath Volume	0.000	14:49	
Air Blank	0.000	14:50	
Control Test	0.080	14:50	
Air Blank	0.000	14:51	
Diagnostics Check	OK	14:51	

Data Download Date/Time: 06/12/2017 15:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003197

INSTRUMENT REGISTERED TO: FWC	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 477627
BREATH TEST OPERATOR: MALTAIS,ANDREW,W	DRY GAS EXPIRY DATE: 06/29/2019
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)	VIOLATION CODE: BUI
ARRESTING OFFICER: MALTAIS,ANDREW	ARREST TIME: 15:00
ARRESTING OFFICER AGENCY: FFWCC (Panama City)	GENDER: Male
LAST AGENCY INSPECTION: 05/11/2017	OBSERVATION PERIOD BEGAN: 15:05

SUBJECT'S LAST NAME: ANTHONY	DATE/TIME OF TEST: 06/03/2017 15:29:35
------------------------------	--

Diagnostics Check	OK	15:26	
Air Blank	0.000	15:27	
Control Test	0.080	15:27	
Air Blank	0.000	15:28	
Subject Sample	# 1	15:28	Subject Test Refused
Breath Volume	0.000	15:28	
Air Blank	0.000	15:28	
Control Test	0.081	15:29	
Air Blank	0.000	15:29	
Diagnostics Check	OK	15:29	

Data Download Date/Time: 06/12/2017 15:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003197

INSTRUMENT REGISTERED TO: FWC

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HANEY,MICHAEL,P

BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)

ARRESTING OFFICER: MOBLEY,BRITTANY

ARRESTING OFFICER AGENCY: FFWCC (Panama City)

LAST AGENCY INSPECTION: 05/11/2017

Date or Time Changed: No

DRY GAS STANDARD LOT #: 477627

DRY GAS EXPIRY DATE: 06/29/2019

VIOLATION CODE: BUI

ARREST TIME: 16:33

GENDER: Female

OBSERVATION PERIOD BEGAN: 16:30

SUBJECT'S LAST NAME: BROOKS

DATE/TIME OF TEST: 06/03/2017 17:10:22

Diagnostics Check	OK	17:05	
Air Blank	0.000	17:05	
Control Test	0.081	17:06	
Air Blank	0.000	17:06	
Subject Sample	# 1	17:08	Subject Test Refused
Breath Volume	0.066	17:08	
Air Blank	0.000	17:09	
Control Test	0.081	17:09	
Air Blank	0.000	17:10	
Diagnostics Check	OK	17:10	

Data Download Date/Time: 06/12/2017 15:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003197

INSTRUMENT REGISTERED TO: FWC	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 477627
BREATH TEST OPERATOR: RAMOS,ROBERT,G	DRY GAS EXPIRY DATE: 06/29/2019
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)	VIOLATION CODE: BUI
ARRESTING OFFICER: RAMOS,ROBERT	ARREST TIME: 16:50
ARRESTING OFFICER AGENCY: FFWCC (Panama City)	GENDER: Male
LAST AGENCY INSPECTION: 05/11/2017	OBSERVATION PERIOD BEGAN: 16:50

SUBJECT'S LAST NAME: PALMER DATE/TIME OF TEST: 06/03/2017 18:08:09

Diagnostics Check	OK	18:05	
Air Blank	0.000	18:05	
Control Test	0.080	18:05	
Air Blank	0.000	18:06	
Subject Sample	# 1	18:06	Subject Test Refused
Breath Volume	0.000	18:06	
Air Blank	0.000	18:07	
Control Test	0.080	18:07	
Air Blank	0.000	18:08	
Diagnostics Check	OK	18:08	

Data Download Date/Time: 06/12/2017 15:56

Subject Test Electronic Data

DATE/TIME OF TEST: 06/17/2017 22:54:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003347

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANG,JACQUES,A
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: DAVENPORT,JOSHUA
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/13/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 24515080A2
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:47

SUBJECT'S LAST NAME: ROCHE VAZQUEZ

DATE/TIME OF TEST: 06/15/2017 00:02:03

Diagnostics Check	OK	23:55
Air Blank	0.000	23:56
Control Test	0.078	23:56
Air Blank	0.000	23:56
Subject Sample # 2	0.191	23:57
Breath Volume	1.621	23:57
Air Blank	0.000	23:58
Air Blank	0.000	23:59
Subject Sample # 1	0.188	00:00
Breath Volume	1.707	00:00
Air Blank	0.000	00:01
Control Test	0.077	00:01
Air Blank	0.000	00:01
Diagnostics Check	OK	00:02

Data Download Date/Time: 07/26/2017 12:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003347

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,JASON,G
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: BROWN,JASON
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A2
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: MILOS

DATE/TIME OF TEST: 06/15/2017 01:26:39

Diagnostics Check	OK	01:20
Air Blank	0.000	01:20
Control Test	0.077	01:20
Air Blank	0.000	01:21
Subject Sample # 1	0.157	01:22
Breath Volume	2.968	01:22
Air Blank	0.000	01:22
Air Blank	0.000	01:24
Subject Sample # 2	0.153	01:25
Breath Volume	3.160	01:25
Air Blank	0.000	01:25
Control Test	0.077	01:26
Air Blank	0.000	01:26
Diagnostics Check	OK	01:26

Data Download Date/Time: 07/26/2017 12:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003347

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCONNELL,J,W
BREATH TEST OPERATOR AGENCY: New Port Richey PD
ARRESTING OFFICER: MCCONNELL,J
ARRESTING OFFICER AGENCY: New Port Richey PD
LAST AGENCY INSPECTION: 06/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A2
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 15:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:00

SUBJECT'S LAST NAME: LIENARD

DATE/TIME OF TEST: 06/15/2017 18:31:51

Diagnostics Check	OK	18:25
Air Blank	0.000	18:26
Control Test	0.076	18:26
Air Blank	0.000	18:26
Subject Sample # 1	0.193	18:27
Breath Volume	2.285	18:27
Air Blank	0.000	18:28
Air Blank	0.000	18:29
Subject Sample # 2	0.191	18:30
Breath Volume	2.718	18:30
Air Blank	0.000	18:30
Control Test	0.076	18:31
Air Blank	0.000	18:31
Diagnostics Check	OK	18:31

Data Download Date/Time: 07/26/2017 12:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003347

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLYTH,JACOB,A
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: BLYTH,JACOB
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A2
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 19:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:07

SUBJECT'S LAST NAME: KOEZENO

DATE/TIME OF TEST: 06/15/2017 20:33:45

Diagnostics Check	OK	20:27
Air Blank	0.000	20:28
Control Test	0.077	20:28
Air Blank	0.000	20:29
Subject Sample # 1	0.000	20:29
Breath Volume	1.722	20:29
Air Blank	0.000	20:29
Air Blank	0.000	20:31
Subject Sample # 2	0.000	20:32
Breath Volume	2.335	20:32
Air Blank	0.000	20:32
Control Test	0.079	20:33
Air Blank	0.000	20:33
Diagnostics Check	OK	20:33

Data Download Date/Time: 07/26/2017 12:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003347

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NIXON,BARRY,
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: NIXON,BARRY
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 06/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A2
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:41

SUBJECT'S LAST NAME: COUTURE

DATE/TIME OF TEST: 06/17/2017 01:07:35

Diagnostics Check	OK	01:01
Air Blank	0.000	01:02
Control Test	0.078	01:02
Air Blank	0.000	01:02
Subject Sample # 1	0.166	01:03
Breath Volume	1.605	01:03
Air Blank	0.000	01:03
Air Blank	0.000	01:05
Subject Sample # 2	0.161	01:06
Breath Volume	1.656	01:06
Air Blank	0.000	01:06
Control Test	0.078	01:07
Air Blank	0.000	01:07
Diagnostics Check	OK	01:07

Data Download Date/Time: 07/26/2017 12:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003347

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNORR,PETER,J
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: KNORR,PETER
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 06/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A2
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:24

SUBJECT'S LAST NAME: VITELLIO

DATE/TIME OF TEST: 06/18/2017 00:51:51

Diagnostics Check	OK	00:44
Air Blank	0.000	00:45
Control Test	0.078	00:45
Air Blank	0.000	00:46
Subject Sample # 1	0.071	00:47
Breath Volume	3.320	00:47
Air Blank	0.000	00:47
Air Blank	0.000	00:49
Subject Sample # 2	0.058	00:50
Breath Volume	1.820	00:50
Air Blank	0.000	00:50
Control Test	0.078	00:51
Air Blank	0.000	00:51
Diagnostics Check	OK	00:51

Data Download Date/Time: 07/26/2017 12:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003347

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KRAUS,CHRISTOPHER,J
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: KRAUS,CHRISTOPHER
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A2
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 12:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:30

SUBJECT'S LAST NAME: CASTILLO

DATE/TIME OF TEST: 06/19/2017 14:01:54

Diagnostics Check	OK	13:52
Air Blank	0.000	13:53
Control Test	0.078	13:53
Air Blank	0.000	13:54
Subject Sample # 1	0.311	13:54
Breath Volume	2.718	13:54
Air Blank	0.000	13:55
Air Blank	0.000	13:57
Subject Sample # 2	0.290	13:57
Breath Volume	2.976	13:57
Air Blank	0.000	13:58
Air Blank	0.000	13:59
Subject Sample # 3	0.308	14:00
Breath Volume	3.757	14:00
Air Blank	0.000	14:01
Control Test	0.076	14:01
Air Blank	0.000	14:01
Diagnostics Check	OK	14:01

Data Download Date/Time: 07/26/2017 12:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003347

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREEN, WILLIAM, E
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: GREEN, WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A2
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:13

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 06/19/2017 21:45:24

Diagnostics Check	OK	21:38
Air Blank	0.000	21:38
Control Test	0.077	21:39
Air Blank	0.000	21:39
Subject Sample # 1	0.084	21:40
Breath Volume	1.687	21:40
Air Blank	0.000	21:40
Air Blank	0.000	21:42
Subject Sample # 2	0.082	21:43
Breath Volume	1.703	21:43
Air Blank	0.000	21:44
Control Test	0.078	21:44
Air Blank	0.000	21:45
Diagnostics Check	OK	21:45

Data Download Date/Time: 07/26/2017 12:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003347

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNORR,PETER,J
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: BLICKER,EVAN
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 06/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A2
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:09

SUBJECT'S LAST NAME: PRESLAR

DATE/TIME OF TEST: 06/22/2017 00:12:06

Diagnostics Check	OK	23:59	
Air Blank	0.000	00:00	
Control Test	0.076	00:00	
Air Blank	0.000	00:01	
Subject Sample	# 1	00:04	No Sample Provided
Breath Volume	0.000	00:04	
Air Blank	0.000	00:04	
Air Blank	0.000	00:06	
Subject Sample	# 2	0.179	00:07
Breath Volume	3.980	00:07	
Air Blank	0.000	00:07	
Air Blank	0.000	00:09	
Subject Sample	# 3	0.173	00:10
Breath Volume	3.847	00:10	
Air Blank	0.000	00:11	
Control Test	0.075	00:11	
Air Blank	0.000	00:12	
Diagnostics Check	OK	00:12	

Data Download Date/Time: 07/26/2017 12:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003347

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNORR,PETER,J
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: KNORR,PETER
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 06/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A2
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: KINCAID

DATE/TIME OF TEST: 06/23/2017 01:22:10

Diagnostics Check	OK	01:15
Air Blank	0.000	01:15
Control Test	0.078	01:16
Air Blank	0.000	01:16
Subject Sample # 1	0.148	01:17
Breath Volume	1.898	01:17
Air Blank	0.000	01:18
Air Blank	0.000	01:19
Subject Sample # 2	0.146	01:20
Breath Volume	2.925	01:20
Air Blank	0.000	01:21
Control Test	0.077	01:21
Air Blank	0.000	01:22
Diagnostics Check	OK	01:22

Data Download Date/Time: 07/26/2017 12:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003347

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALVARADO,ANGEL,L
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: BLICKER,EVAN
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 06/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A2
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:05

SUBJECT'S LAST NAME: PARKER

DATE/TIME OF TEST: 06/23/2017 05:32:51

Diagnostics Check	OK	05:26
Air Blank	0.000	05:26
Control Test	0.077	05:26
Air Blank	0.000	05:27
Subject Sample # 1	0.262	05:28
Breath Volume	2.050	05:28
Air Blank	0.000	05:28
Air Blank	0.000	05:30
Subject Sample # 2	0.245	05:31
Breath Volume	2.308	05:31
Air Blank	0.000	05:31
Control Test	0.077	05:32
Air Blank	0.000	05:32
Diagnostics Check	OK	05:32

Data Download Date/Time: 07/26/2017 12:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003347

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNORR,PETER,J
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: FITCHEN,BRIAN
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 06/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A2
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: MICCI

DATE/TIME OF TEST: 06/24/2017 01:44:55

Diagnostics Check	OK	01:38
Air Blank	0.000	01:38
Control Test	0.077	01:39
Air Blank	0.000	01:39
Subject Sample # 1	0.198	01:40
Breath Volume	2.386	01:40
Air Blank	0.000	01:40
Air Blank	0.000	01:42
Subject Sample # 2	0.201	01:43
Breath Volume	2.351	01:43
Air Blank	0.000	01:44
Control Test	0.077	01:44
Air Blank	0.000	01:44
Diagnostics Check	OK	01:44

Data Download Date/Time: 07/26/2017 12:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003347

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NOTTINGHAM,CHRISTOPHER,R
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: NOTTINGHAM,CHRISTOPHER
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A2
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 05:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:08

SUBJECT'S LAST NAME: TAWILL

DATE/TIME OF TEST: 06/24/2017 06:35:25

Diagnostics Check	OK	06:29
Air Blank	0.000	06:29
Control Test	0.078	06:29
Air Blank	0.000	06:30
Subject Sample # 1	0.000	06:30
Breath Volume	3.164	06:30
Air Blank	0.000	06:31
Air Blank	0.000	06:33
Subject Sample # 2	0.000	06:34
Breath Volume	3.187	06:34
Air Blank	0.000	06:34
Control Test	0.079	06:34
Air Blank	0.000	06:35
Diagnostics Check	OK	06:35

Data Download Date/Time: 07/26/2017 12:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003347

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,CLINT,R
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)
ARRESTING OFFICER: WILLIAMS,CLINT
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 06/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A2
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: BUI
ARREST TIME: 19:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: BURTON

DATE/TIME OF TEST: 06/24/2017 20:32:19

Diagnostics Check	OK	20:20	
Air Blank	0.000	20:21	
Control Test	0.078	20:21	
Air Blank	0.000	20:22	
Subject Sample	# 1	20:25	No Sample Provided
Breath Volume	0.000	20:25	
Air Blank	0.000	20:25	
Air Blank	0.000	20:27	
Subject Sample	# 2	20:30	No Sample Provided
Breath Volume	0.000	20:30	
Air Blank	0.000	20:31	
Control Test	0.078	20:31	
Air Blank	0.000	20:32	
Diagnostics Check	OK	20:32	

Data Download Date/Time: 07/26/2017 12:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003347

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORRIS, ADAM, M
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: MORRIS, ADAM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A2
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: TYLER

DATE/TIME OF TEST: 06/25/2017 01:48:17

Diagnostics Check	OK	01:40
Air Blank	0.000	01:41
Control Test	0.078	01:41
Air Blank	0.000	01:42
Subject Sample # 1	0.130	01:43
Breath Volume	2.445	01:43
Air Blank	0.000	01:44
Air Blank	0.000	01:46
Subject Sample # 2	0.130	01:46
Breath Volume	1.804	01:46
Air Blank	0.000	01:47
Control Test	0.078	01:47
Air Blank	0.000	01:48
Diagnostics Check	OK	01:48

Data Download Date/Time: 07/26/2017 12:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003347

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLYTH,JACOB,A
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: BLYTH,JACOB
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A2
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:33

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 06/25/2017 03:00:29

Diagnostics Check	OK	02:54
Air Blank	0.000	02:54
Control Test	0.077	02:54
Air Blank	0.000	02:55
Subject Sample # 1	0.213	02:55
Breath Volume	4.308	02:55
Air Blank	0.000	02:56
Air Blank	0.000	02:58
Subject Sample # 2	0.209	02:58
Breath Volume	4.417	02:58
Air Blank	0.000	02:59
Control Test	0.077	02:59
Air Blank	0.000	03:00
Diagnostics Check	OK	03:00

Data Download Date/Time: 07/26/2017 12:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003347

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NOTTINGHAM,CHRISTOPHER,R
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: NOTTINGHAM,CHRISTOPHER
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A2
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:15

SUBJECT'S LAST NAME: KOHLER

DATE/TIME OF TEST: 06/25/2017 22:42:11

Diagnostics Check	OK	22:36
Air Blank	0.000	22:36
Control Test	0.078	22:36
Air Blank	0.000	22:37
Subject Sample # 1	0.120	22:37
Breath Volume	1.164	22:37
Air Blank	0.000	22:38
Air Blank	0.000	22:40
Subject Sample # 2	0.117	22:40
Breath Volume	2.390	22:40
Air Blank	0.000	22:41
Control Test	0.078	22:41
Air Blank	0.000	22:42
Diagnostics Check	OK	22:42

Data Download Date/Time: 07/26/2017 12:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003347

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NOTTINGHAM,CHRISTOPHER,R
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: NOTTINGHAM,CHRISTOPHER
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A2
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:25

SUBJECT'S LAST NAME: BOWMAN

DATE/TIME OF TEST: 06/26/2017 04:52:46

Diagnostics Check	OK	04:46
Air Blank	0.000	04:46
Control Test	0.077	04:46
Air Blank	0.000	04:47
Subject Sample # 1	0.258	04:47
Breath Volume	3.113	04:47
Air Blank	0.000	04:48
Air Blank	0.000	04:50
Subject Sample # 2	0.259	04:51
Breath Volume	2.683	04:51
Air Blank	0.000	04:51
Control Test	0.078	04:52
Air Blank	0.000	04:52
Diagnostics Check	OK	04:52

Data Download Date/Time: 07/26/2017 12:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003347

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: IVEY,BRIAN,J
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: IVEY,BRIAN
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 06/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A2
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:22

SUBJECT'S LAST NAME: SCHIPPAN

DATE/TIME OF TEST: 06/27/2017 02:51:58

Diagnostics Check	OK	02:43
Air Blank	0.000	02:43
Control Test	0.079	02:44
Air Blank	0.000	02:44
Subject Sample # 1	0.240	02:45
Breath Volume	1.457	02:45
Air Blank	0.000	02:46
Air Blank	0.000	02:48
Subject Sample # 2	0.229	02:50
Breath Volume	1.246	02:50
Air Blank	0.000	02:51
Control Test	0.078	02:51
Air Blank	0.000	02:51
Diagnostics Check	OK	02:51

Data Download Date/Time: 07/26/2017 12:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003347

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREEN, WILLIAM, E
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: GREEN, WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A2
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:28

SUBJECT'S LAST NAME: NEWMAN

DATE/TIME OF TEST: 06/27/2017 04:01:36

Diagnostics Check	OK	03:55
Air Blank	0.000	03:55
Control Test	0.078	03:55
Air Blank	0.000	03:56
Subject Sample # 1	0.000	03:57
Breath Volume	3.417	03:57
Air Blank	0.000	03:57
Air Blank	0.000	03:59
Subject Sample # 2	0.000	04:00
Breath Volume	3.691	04:00
Air Blank	0.000	04:00
Control Test	0.079	04:01
Air Blank	0.000	04:01
Diagnostics Check	OK	04:01

Data Download Date/Time: 07/26/2017 12:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003347

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREEN, WILLIAM, E
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: GREEN, WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A2
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:15

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 06/28/2017 04:43:32

Diagnostics Check	OK	04:37
Air Blank	0.000	04:37
Control Test	0.078	04:37
Air Blank	0.000	04:38
Subject Sample # 1	0.161	04:38
Breath Volume	1.222	04:38
Air Blank	0.000	04:39
Air Blank	0.000	04:41
Subject Sample # 2	0.159	04:42
Breath Volume	1.222	04:42
Air Blank	0.000	04:42
Control Test	0.077	04:42
Air Blank	0.000	04:43
Diagnostics Check	OK	04:43

Data Download Date/Time: 07/26/2017 12:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003347

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NOTTINGHAM,CHRISTOPHER,R
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: NOTTINGHAM,CHRISTOPHER
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A2
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: CRIBBS

DATE/TIME OF TEST: 06/29/2017 01:22:27

Diagnostics Check	OK	01:16
Air Blank	0.000	01:16
Control Test	0.078	01:17
Air Blank	0.000	01:17
Subject Sample # 1	0.000	01:18
Breath Volume	1.933	01:18
Air Blank	0.000	01:18
Air Blank	0.000	01:20
Subject Sample # 2	0.000	01:21
Breath Volume	1.828	01:21
Air Blank	0.000	01:21
Control Test	0.078	01:21
Air Blank	0.000	01:22
Diagnostics Check	OK	01:22

Data Download Date/Time: 07/26/2017 12:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003347

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NIXON,BARRY,
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: NIXON,BARRY
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 06/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A2
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:22

SUBJECT'S LAST NAME: RUSSO

DATE/TIME OF TEST: 06/29/2017 23:49:04

Diagnostics Check	OK	23:43
Air Blank	0.000	23:43
Control Test	0.079	23:43
Air Blank	0.000	23:44
Subject Sample # 1	0.000	23:44
Breath Volume	2.921	23:44
Air Blank	0.000	23:45
Air Blank	0.000	23:47
Subject Sample # 2	0.000	23:47
Breath Volume	2.183	23:47
Air Blank	0.000	23:48
Control Test	0.077	23:48
Air Blank	0.000	23:48
Diagnostics Check	OK	23:49

Data Download Date/Time: 07/26/2017 12:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003347

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANGELICCHI,JOSEPH,S
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: LANE,M
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 06/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A2
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:54

SUBJECT'S LAST NAME: MONSALVE

DATE/TIME OF TEST: 06/30/2017 03:24:29

Diagnostics Check	OK	03:18
Air Blank	0.000	03:18
Control Test	0.078	03:18
Air Blank	0.000	03:19
Subject Sample # 1	0.083	03:20
Breath Volume	3.425	03:20
Air Blank	0.000	03:20
Air Blank	0.000	03:22
Subject Sample # 2	0.081	03:23
Breath Volume	3.632	03:23
Air Blank	0.000	03:23
Control Test	0.078	03:23
Air Blank	0.000	03:24
Diagnostics Check	OK	03:24

Data Download Date/Time: 07/26/2017 12:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003347

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORRIS, ADAM, M
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: MORRIS, ADAM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A2
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 06/30/2017 04:56:32

Diagnostics Check	OK	04:50
Air Blank	0.000	04:50
Control Test	0.078	04:51
Air Blank	0.000	04:51
Subject Sample # 1	0.149	04:52
Breath Volume	4.375	04:52
Air Blank	0.000	04:52
Air Blank	0.000	04:54
Subject Sample # 2	0.145	04:55
Breath Volume	3.898	04:55
Air Blank	0.000	04:55
Control Test	0.077	04:55
Air Blank	0.000	04:56
Diagnostics Check	OK	04:56

Data Download Date/Time: 07/26/2017 12:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003347

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCARTHY,KEITH,
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: MURPHY,J
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 06/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A2
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 10:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:30

SUBJECT'S LAST NAME: O'DONNELL

DATE/TIME OF TEST: 06/30/2017 11:17:12

Diagnostics Check	OK	11:10
Air Blank	0.000	11:11
Control Test	0.078	11:11
Air Blank	0.000	11:12
Subject Sample # 1	0.098	11:12
Breath Volume	3.199	11:12
Air Blank	0.000	11:13
Air Blank	0.000	11:15
Subject Sample # 2	0.089	11:15
Breath Volume	2.324	11:15
Air Blank	0.000	11:16
Control Test	0.079	11:16
Air Blank	0.000	11:17
Diagnostics Check	OK	11:17

Data Download Date/Time: 07/26/2017 12:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: COSTA,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:14

SUBJECT'S LAST NAME: FELS

DATE/TIME OF TEST: 06/09/2017 04:40:40

Diagnostics Check	OK	04:34
Air Blank	0.000	04:34
Control Test	0.079	04:35
Air Blank	0.000	04:35
Subject Sample # 1	0.000	04:36
Breath Volume	2.246	04:36
Air Blank	0.000	04:36
Air Blank	0.000	04:38
Subject Sample # 2	0.000	04:39
Breath Volume	2.558	04:39
Air Blank	0.000	04:39
Control Test	0.079	04:40
Air Blank	0.000	04:40
Diagnostics Check	OK	04:40

Data Download Date/Time: 07/11/2017 18:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK,CASEY,S
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: CAMPBELL,J
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:39

SUBJECT'S LAST NAME: BRIGGS

DATE/TIME OF TEST: 06/10/2017 04:06:26

Diagnostics Check	OK	03:59
Air Blank	0.000	04:00
Control Test	0.080	04:00
Air Blank	0.000	04:00
Subject Sample # 1	0.098	04:01
Breath Volume	2.511	04:01
Air Blank	0.000	04:02
Air Blank	0.000	04:04
Subject Sample # 2	0.097	04:04
Breath Volume	3.042	04:04
Air Blank	0.000	04:05
Control Test	0.079	04:05
Air Blank	0.000	04:06
Diagnostics Check	OK	04:06

Data Download Date/Time: 07/11/2017 18:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:10

SUBJECT'S LAST NAME: FRANQUI CARRERO

DATE/TIME OF TEST: 06/10/2017 05:39:38

Diagnostics Check	OK	05:30
Air Blank	0.000	05:31
Control Test	0.080	05:31
Air Blank	0.000	05:31
Subject Sample # 1	0.190	05:34
Breath Volume	1.695	05:34
Air Blank	0.000	05:35
Air Blank	0.000	05:36
Subject Sample # 2	0.191	05:38
Breath Volume	1.378	05:38
Air Blank	0.000	05:38
Control Test	0.078	05:39
Air Blank	0.000	05:39
Diagnostics Check	OK	05:39

Data Download Date/Time: 07/11/2017 18:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BLANCHARD, J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:43

SUBJECT'S LAST NAME: RAMIREZ

DATE/TIME OF TEST: 06/10/2017 21:17:16

Diagnostics Check	OK	21:03	
Air Blank	0.000	21:03	
Control Test	0.079	21:04	
Air Blank	0.000	21:04	
Subject Sample	# 1 0.167	21:07	Volume Not Met
Breath Volume	0.769	21:07	
Air Blank	0.000	21:08	
Air Blank	0.000	21:10	
Subject Sample	# 2 0.197	21:11	
Breath Volume	1.464	21:11	
Air Blank	0.000	21:12	
Air Blank	0.000	21:14	
Subject Sample	# 3 0.195	21:15	
Breath Volume	1.324	21:15	
Air Blank	0.000	21:16	
Control Test	0.077	21:16	
Air Blank	0.000	21:17	
Diagnostics Check	OK	21:17	

Data Download Date/Time: 07/11/2017 18:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:04

SUBJECT'S LAST NAME: GARCIA ROMAN

DATE/TIME OF TEST: 06/11/2017 01:31:42

Diagnostics Check	OK	01:25
Air Blank	0.000	01:25
Control Test	0.078	01:25
Air Blank	0.000	01:26
Subject Sample # 1	0.137	01:27
Breath Volume	2.648	01:27
Air Blank	0.000	01:27
Air Blank	0.000	01:29
Subject Sample # 2	0.134	01:30
Breath Volume	2.429	01:30
Air Blank	0.000	01:30
Control Test	0.079	01:31
Air Blank	0.000	01:31
Diagnostics Check	OK	01:31

Data Download Date/Time: 07/11/2017 18:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DE JESUS, V
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: CASTANEDA MOREJON

DATE/TIME OF TEST: 06/11/2017 03:43:21

Diagnostics Check	OK	03:35
Air Blank	0.000	03:35
Control Test	0.079	03:36
Air Blank	0.000	03:36
Subject Sample # 1	0.153	03:37
Breath Volume	1.371	03:37
Air Blank	0.000	03:37
Air Blank	0.000	03:39
Subject Sample # 2	0.151	03:41
Breath Volume	1.328	03:41
Air Blank	0.000	03:42
Control Test	0.079	03:42
Air Blank	0.000	03:43
Diagnostics Check	OK	03:43

Data Download Date/Time: 07/11/2017 18:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PECYLAK, M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:22

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 06/12/2017 02:49:28

Diagnostics Check	OK	02:42
Air Blank	0.000	02:43
Control Test	0.079	02:43
Air Blank	0.000	02:44
Subject Sample # 1	0.130	02:44
Breath Volume	1.585	02:44
Air Blank	0.000	02:45
Air Blank	0.000	02:47
Subject Sample # 2	0.127	02:47
Breath Volume	1.781	02:47
Air Blank	0.000	02:48
Control Test	0.079	02:48
Air Blank	0.000	02:49
Diagnostics Check	OK	02:49

Data Download Date/Time: 07/11/2017 18:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PECYLAK,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:39

SUBJECT'S LAST NAME: MAURY

DATE/TIME OF TEST: 06/13/2017 03:07:06

Diagnostics Check	OK	03:00
Air Blank	0.000	03:00
Control Test	0.079	03:01
Air Blank	0.000	03:01
Subject Sample # 1	0.174	03:02
Breath Volume	1.238	03:02
Air Blank	0.000	03:03
Air Blank	0.000	03:04
Subject Sample # 2	0.191	03:05
Breath Volume	1.894	03:05
Air Blank	0.000	03:06
Control Test	0.077	03:06
Air Blank	0.000	03:07
Diagnostics Check	OK	03:07

Data Download Date/Time: 07/11/2017 18:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARLAUG,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:54

SUBJECT'S LAST NAME: SLATTON

DATE/TIME OF TEST: 06/15/2017 03:20:39

Diagnostics Check	OK	03:14
Air Blank	0.000	03:14
Control Test	0.079	03:15
Air Blank	0.000	03:15
Subject Sample # 1	0.116	03:16
Breath Volume	2.015	03:16
Air Blank	0.000	03:16
Air Blank	0.000	03:18
Subject Sample # 2	0.114	03:19
Breath Volume	2.769	03:19
Air Blank	0.000	03:19
Control Test	0.078	03:20
Air Blank	0.000	03:20
Diagnostics Check	OK	03:20

Data Download Date/Time: 07/11/2017 18:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK,CASEY,S
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRONSON,R
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:16

SUBJECT'S LAST NAME: HOUSLEY

DATE/TIME OF TEST: 06/15/2017 04:47:57

Diagnostics Check	OK	04:36	
Air Blank	0.000	04:36	
Control Test	0.079	04:37	
Air Blank	0.000	04:37	
Subject Sample	# 1 0.190	04:40	Volume Not Met
Breath Volume	0.957	04:40	
Air Blank	0.000	04:41	
Air Blank	0.000	04:43	
Subject Sample	# 2 0.160	04:46	Volume Not Met
Breath Volume	0.046	04:46	
Air Blank	0.000	04:46	
Control Test	0.079	04:47	
Air Blank	0.000	04:47	
Diagnostics Check	OK	04:47	

Data Download Date/Time: 07/11/2017 18:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FOWLER, J
ARRESTING OFFICER AGENCY: Plant City PD
LAST AGENCY INSPECTION: 06/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:56

SUBJECT'S LAST NAME: LEE

DATE/TIME OF TEST: 06/16/2017 00:23:44

Diagnostics Check	OK	00:16
Air Blank	0.000	00:17
Control Test	0.079	00:17
Air Blank	0.000	00:18
Subject Sample # 1	0.177	00:18
Breath Volume	2.593	00:18
Air Blank	0.000	00:19
Air Blank	0.000	00:21
Subject Sample # 2	0.176	00:22
Breath Volume	2.652	00:22
Air Blank	0.000	00:22
Control Test	0.079	00:23
Air Blank	0.000	00:23
Diagnostics Check	OK	00:23

Data Download Date/Time: 07/11/2017 18:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ORNDORFF,W
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:37

SUBJECT'S LAST NAME: CHEEVES

DATE/TIME OF TEST: 06/17/2017 01:05:23

Diagnostics Check	OK	00:57
Air Blank	0.000	00:58
Control Test	0.079	00:58
Air Blank	0.000	00:59
Subject Sample # 1	0.219	01:00
Breath Volume	1.500	01:00
Air Blank	0.000	01:01
Air Blank	0.000	01:03
Subject Sample # 2	0.206	01:03
Breath Volume	1.925	01:03
Air Blank	0.000	01:04
Control Test	0.077	01:04
Air Blank	0.000	01:05
Diagnostics Check	OK	01:05

Data Download Date/Time: 07/11/2017 18:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK,CASEY,S
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ORNDORFF,W
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:06

SUBJECT'S LAST NAME: MUSIAL

DATE/TIME OF TEST: 06/18/2017 03:33:23

Diagnostics Check	OK	03:26
Air Blank	0.000	03:27
Control Test	0.080	03:27
Air Blank	0.000	03:28
Subject Sample # 1	0.246	03:28
Breath Volume	3.207	03:28
Air Blank	0.000	03:29
Air Blank	0.000	03:31
Subject Sample # 2	0.247	03:31
Breath Volume	2.777	03:31
Air Blank	0.000	03:32
Control Test	0.078	03:32
Air Blank	0.000	03:33
Diagnostics Check	OK	03:33

Data Download Date/Time: 07/11/2017 18:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARLAUG,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:33

SUBJECT'S LAST NAME: DELEON

DATE/TIME OF TEST: 06/18/2017 05:00:03

Diagnostics Check	OK	04:53
Air Blank	0.000	04:53
Control Test	0.079	04:54
Air Blank	0.000	04:54
Subject Sample # 1	0.129	04:55
Breath Volume	2.332	04:55
Air Blank	0.000	04:56
Air Blank	0.000	04:57
Subject Sample # 2	0.125	04:58
Breath Volume	2.308	04:58
Air Blank	0.000	04:59
Control Test	0.078	04:59
Air Blank	0.000	04:59
Diagnostics Check	OK	05:00

Data Download Date/Time: 07/11/2017 18:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GONZALEZ,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:42

SUBJECT'S LAST NAME: LANHAM

DATE/TIME OF TEST: 06/20/2017 03:10:24

Diagnostics Check	OK	03:02
Air Blank	0.000	03:02
Control Test	0.080	03:03
Air Blank	0.000	03:03
Subject Sample # 1	0.149	03:05
Breath Volume	1.828	03:05
Air Blank	0.000	03:06
Air Blank	0.000	03:08
Subject Sample # 2	0.145	03:08
Breath Volume	1.476	03:08
Air Blank	0.000	03:09
Control Test	0.078	03:09
Air Blank	0.000	03:10
Diagnostics Check	OK	03:10

Data Download Date/Time: 07/11/2017 18:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LARA,A
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:41

SUBJECT'S LAST NAME: HORTON

DATE/TIME OF TEST: 06/21/2017 23:08:01

Diagnostics Check	OK	23:01
Air Blank	0.000	23:02
Control Test	0.079	23:02
Air Blank	0.000	23:03
Subject Sample # 1	0.000	23:03
Breath Volume	1.691	23:03
Air Blank	0.000	23:04
Air Blank	0.000	23:06
Subject Sample # 2	0.000	23:06
Breath Volume	2.277	23:06
Air Blank	0.000	23:07
Control Test	0.079	23:07
Air Blank	0.000	23:07
Diagnostics Check	OK	23:08

Data Download Date/Time: 07/11/2017 18:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FRANKLIN,R
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:44

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 06/24/2017 01:15:29

Diagnostics Check	OK	01:04	
Air Blank	0.000	01:04	
Control Test	0.079	01:04	
Air Blank	0.000	01:05	
Subject Sample	# 1	01:08	No Sample Provided
Breath Volume	0.000	01:08	
Air Blank	0.000	01:09	
Air Blank	0.000	01:10	
Subject Sample	# 2	01:14	No Sample Provided
Breath Volume	0.000	01:14	
Air Blank	0.000	01:14	
Control Test	0.079	01:14	
Air Blank	0.000	01:15	
Diagnostics Check	OK	01:15	

Data Download Date/Time: 07/11/2017 18:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FRANKLIN,R
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 06/24/2017 01:51:55

Diagnostics Check	OK	01:39	
Air Blank	0.000	01:40	
Control Test	0.078	01:40	
Air Blank	0.000	01:40	
Subject Sample	# 1	01:44	No Sample Provided
Breath Volume	0.000	01:44	
Air Blank	0.000	01:44	
Air Blank	0.000	01:46	
Subject Sample	# 2	0.124 01:47	
Breath Volume	1.421	01:47	
Air Blank	0.000	01:47	
Air Blank	0.000	01:49	
Subject Sample	# 3	0.118 01:50	
Breath Volume	1.695	01:50	
Air Blank	0.000	01:51	
Control Test	0.078	01:51	
Air Blank	0.000	01:51	
Diagnostics Check	OK	01:51	

Data Download Date/Time: 07/11/2017 18:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRIMBLE,TERESA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRONSON,R
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:16

SUBJECT'S LAST NAME: GUZMAN

DATE/TIME OF TEST: 06/24/2017 23:43:24

Diagnostics Check	OK	23:36
Air Blank	0.000	23:36
Control Test	0.080	23:37
Air Blank	0.000	23:37
Subject Sample # 1	0.179	23:38
Breath Volume	1.757	23:38
Air Blank	0.000	23:38
Air Blank	0.000	23:40
Subject Sample # 2	0.175	23:41
Breath Volume	2.128	23:41
Air Blank	0.000	23:42
Control Test	0.079	23:42
Air Blank	0.000	23:43
Diagnostics Check	OK	23:43

Data Download Date/Time: 07/11/2017 18:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRIMBLE,TERESA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LYON,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:24

SUBJECT'S LAST NAME: CONLEY

DATE/TIME OF TEST: 06/24/2017 23:54:28

Diagnostics Check	OK	23:47
Air Blank	0.000	23:47
Control Test	0.078	23:47
Air Blank	0.000	23:48
Subject Sample # 1	0.231	23:49
Breath Volume	1.812	23:49
Air Blank	0.000	23:50
Air Blank	0.000	23:52
Subject Sample # 2	0.213	23:52
Breath Volume	1.246	23:52
Air Blank	0.000	23:53
Control Test	0.077	23:53
Air Blank	0.000	23:54
Diagnostics Check	OK	23:54

Data Download Date/Time: 07/11/2017 18:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS, S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: KUMAR

DATE/TIME OF TEST: 06/25/2017 02:21:26

Diagnostics Check	OK	02:15
Air Blank	0.000	02:15
Control Test	0.078	02:16
Air Blank	0.000	02:16
Subject Sample # 1	0.219	02:17
Breath Volume	1.636	02:17
Air Blank	0.000	02:17
Air Blank	0.000	02:19
Subject Sample # 2	0.212	02:19
Breath Volume	1.757	02:19
Air Blank	0.000	02:20
Control Test	0.078	02:20
Air Blank	0.000	02:21
Diagnostics Check	OK	02:21

Data Download Date/Time: 07/11/2017 18:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VAN TREESE,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:43

SUBJECT'S LAST NAME: SILVA REYES

DATE/TIME OF TEST: 06/25/2017 05:11:00

Diagnostics Check	OK	05:03
Air Blank	0.000	05:04
Control Test	0.079	05:04
Air Blank	0.000	05:05
Subject Sample # 1	0.187	05:06
Breath Volume	1.335	05:06
Air Blank	0.000	05:06
Air Blank	0.000	05:08
Subject Sample # 2	0.191	05:09
Breath Volume	2.441	05:09
Air Blank	0.000	05:10
Control Test	0.078	05:10
Air Blank	0.000	05:10
Diagnostics Check	OK	05:11

Data Download Date/Time: 07/11/2017 18:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRIMBLE,TERESA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SULLIVAN,T
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 05:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:07

SUBJECT'S LAST NAME: GUEVARA

DATE/TIME OF TEST: 06/25/2017 06:34:48

Diagnostics Check	OK	06:28
Air Blank	0.000	06:28
Control Test	0.079	06:29
Air Blank	0.000	06:29
Subject Sample # 1	0.112	06:30
Breath Volume	3.835	06:30
Air Blank	0.000	06:30
Air Blank	0.000	06:32
Subject Sample # 2	0.108	06:33
Breath Volume	2.437	06:33
Air Blank	0.000	06:33
Control Test	0.078	06:34
Air Blank	0.000	06:34
Diagnostics Check	OK	06:34

Data Download Date/Time: 07/11/2017 18:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRABAND, M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:15

SUBJECT'S LAST NAME: PUPELLO

DATE/TIME OF TEST: 06/25/2017 22:41:26

Diagnostics Check	OK	22:35
Air Blank	0.000	22:36
Control Test	0.079	22:36
Air Blank	0.000	22:36
Subject Sample # 1	0.000	22:37
Breath Volume	1.441	22:37
Air Blank	0.000	22:37
Air Blank	0.000	22:39
Subject Sample # 2	0.000	22:40
Breath Volume	2.195	22:40
Air Blank	0.000	22:40
Control Test	0.079	22:40
Air Blank	0.000	22:41
Diagnostics Check	OK	22:41

Data Download Date/Time: 07/11/2017 18:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRIMBLE,TERESA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DAWSY,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:32

SUBJECT'S LAST NAME: MOWERY

DATE/TIME OF TEST: 06/25/2017 22:59:18

Diagnostics Check	OK	22:53
Air Blank	0.000	22:53
Control Test	0.080	22:53
Air Blank	0.000	22:54
Subject Sample # 1	0.175	22:54
Breath Volume	3.519	22:54
Air Blank	0.000	22:55
Air Blank	0.000	22:57
Subject Sample # 2	0.175	22:57
Breath Volume	3.437	22:57
Air Blank	0.000	22:58
Control Test	0.078	22:58
Air Blank	0.000	22:59
Diagnostics Check	OK	22:59

Data Download Date/Time: 07/11/2017 18:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MILLER,J
ARRESTING OFFICER AGENCY: Temple Terrace PD
LAST AGENCY INSPECTION: 06/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:42

SUBJECT'S LAST NAME: HOUGH

DATE/TIME OF TEST: 06/26/2017 05:11:09

Diagnostics Check	OK	05:03
Air Blank	0.000	05:03
Control Test	0.079	05:04
Air Blank	0.000	05:04
Subject Sample # 1	0.166	05:05
Breath Volume	1.597	05:05
Air Blank	0.000	05:06
Air Blank	0.000	05:07
Subject Sample # 2	0.157	05:09
Breath Volume	1.902	05:09
Air Blank	0.000	05:10
Control Test	0.079	05:10
Air Blank	0.000	05:11
Diagnostics Check	OK	05:11

Data Download Date/Time: 07/11/2017 18:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DUNCAN,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:12

SUBJECT'S LAST NAME: MCRAE

DATE/TIME OF TEST: 06/27/2017 00:38:40

Diagnostics Check	OK	00:32
Air Blank	0.000	00:32
Control Test	0.079	00:33
Air Blank	0.000	00:33
Subject Sample # 1	0.076	00:34
Breath Volume	1.804	00:34
Air Blank	0.000	00:34
Air Blank	0.000	00:36
Subject Sample # 2	0.075	00:37
Breath Volume	1.230	00:37
Air Blank	0.000	00:37
Control Test	0.079	00:38
Air Blank	0.000	00:38
Diagnostics Check	OK	00:38

Data Download Date/Time: 07/11/2017 18:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRIMBLE,TERESA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BLANCHARD,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:41

SUBJECT'S LAST NAME: HILL

DATE/TIME OF TEST: 06/27/2017 01:13:52

Diagnostics Check	OK	01:01	
Air Blank	0.000	01:01	
Control Test	0.079	01:02	
Air Blank	0.000	01:02	
Subject Sample	# 1 0.204	01:05	Volume Not Met
Breath Volume	0.625	01:05	
Air Blank	0.000	01:06	
Air Blank	0.000	01:08	
Subject Sample	# 2 0.206	01:08	
Breath Volume	2.378	01:08	
Air Blank	0.000	01:09	
Air Blank	0.000	01:11	
Subject Sample	# 3 0.195	01:12	
Breath Volume	2.019	01:12	
Air Blank	0.000	01:12	
Control Test	0.077	01:13	
Air Blank	0.000	01:13	
Diagnostics Check	OK	01:13	

Data Download Date/Time: 07/11/2017 18:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:37

SUBJECT'S LAST NAME: LEWIS

DATE/TIME OF TEST: 06/28/2017 04:08:27

Diagnostics Check	OK	03:57	
Air Blank	0.000	03:57	
Control Test	0.080	03:58	
Air Blank	0.000	03:58	
Subject Sample	# 1 0.182	04:01	Volume Not Met
Breath Volume	0.210	04:01	
Air Blank	0.000	04:02	
Air Blank	0.000	04:04	
Subject Sample	# 2	04:07	Subject Test Refused
Breath Volume	0.000	04:07	
Air Blank	0.000	04:07	
Control Test	0.078	04:07	
Air Blank	0.000	04:08	
Diagnostics Check	OK	04:08	

Data Download Date/Time: 07/11/2017 18:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:38

SUBJECT'S LAST NAME: SERRANO

DATE/TIME OF TEST: 06/29/2017 03:02:35

Diagnostics Check	OK	02:58	
Air Blank	0.000	02:58	
Control Test	0.079	02:59	
Air Blank	0.000	02:59	
Subject Sample	# 1	03:01	Subject Test Refused
Breath Volume	0.000	03:01	
Air Blank	0.000	03:01	
Control Test	0.081	03:01	
Air Blank	0.000	03:02	
Diagnostics Check	OK	03:02	

Data Download Date/Time: 07/11/2017 18:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRIMBLE,TERESA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MACIAS,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:26

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 06/30/2017 01:58:33

Diagnostics Check	OK	01:46	
Air Blank	0.000	01:47	
Control Test	0.080	01:47	
Air Blank	0.000	01:48	
Subject Sample	# 1 0.132	01:51	Volume Not Met
Breath Volume	0.066	01:51	
Air Blank	0.000	01:51	
Air Blank	0.000	01:53	
Subject Sample	# 2 0.114	01:56	Volume Not Met
Breath Volume	0.054	01:56	
Air Blank	0.000	01:57	
Control Test	0.077	01:57	
Air Blank	0.000	01:58	
Diagnostics Check	OK	01:58	

Data Download Date/Time: 07/11/2017 18:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRIMBLE,TERESA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MACIAS,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:26

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 06/30/2017 02:33:02

Diagnostics Check	OK	02:21	
Air Blank	0.000	02:21	
Control Test	0.079	02:22	
Air Blank	0.000	02:22	
Subject Sample	# 1 0.116	02:25	Volume Not Met
Breath Volume	0.117	02:25	
Air Blank	0.000	02:26	
Air Blank	0.000	02:28	
Subject Sample	# 2 0.118	02:31	Volume Not Met
Breath Volume	0.203	02:31	
Air Blank	0.000	02:32	
Control Test	0.077	02:32	
Air Blank	0.000	02:32	
Diagnostics Check	OK	02:33	

Data Download Date/Time: 07/11/2017 18:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRENNAN, D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:16

SUBJECT'S LAST NAME: MCCREA

DATE/TIME OF TEST: 06/30/2017 03:43:01

Diagnostics Check	OK	03:36
Air Blank	0.000	03:36
Control Test	0.078	03:37
Air Blank	0.000	03:37
Subject Sample # 1	0.000	03:38
Breath Volume	2.136	03:38
Air Blank	0.000	03:39
Air Blank	0.000	03:41
Subject Sample # 2	0.000	03:41
Breath Volume	1.242	03:41
Air Blank	0.000	03:42
Control Test	0.078	03:42
Air Blank	0.000	03:42
Diagnostics Check	OK	03:43

Data Download Date/Time: 07/11/2017 18:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DUNCAN, M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 18:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:44

SUBJECT'S LAST NAME: BARNWELL

DATE/TIME OF TEST: 06/30/2017 20:12:20

Diagnostics Check	OK	20:04
Air Blank	0.000	20:05
Control Test	0.080	20:05
Air Blank	0.000	20:06
Subject Sample # 1	0.235	20:07
Breath Volume	1.453	20:07
Air Blank	0.000	20:08
Air Blank	0.000	20:10
Subject Sample # 2	0.241	20:10
Breath Volume	1.589	20:10
Air Blank	0.000	20:11
Control Test	0.079	20:11
Air Blank	0.000	20:12
Diagnostics Check	OK	20:12

Data Download Date/Time: 07/11/2017 18:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BENNETT,CHAD
ARRESTING OFFICER AGENCY: U of So FL Campus PD
LAST AGENCY INSPECTION: 06/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:09

SUBJECT'S LAST NAME: DEES

DATE/TIME OF TEST: 06/06/2017 02:38:17

Diagnostics Check	OK	02:29
Air Blank	0.000	02:29
Control Test	0.080	02:30
Air Blank	0.000	02:30
Subject Sample # 1	0.172	02:33
Breath Volume	1.742	02:33
Air Blank	0.000	02:34
Air Blank	0.000	02:36
Subject Sample # 2	0.176	02:36
Breath Volume	1.375	02:36
Air Blank	0.000	02:37
Control Test	0.079	02:37
Air Blank	0.000	02:38
Diagnostics Check	OK	02:38

Data Download Date/Time: 07/11/2017 19:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:21

SUBJECT'S LAST NAME: SAYRE

DATE/TIME OF TEST: 06/07/2017 03:44:34

Diagnostics Check	OK	03:42	
Air Blank	0.000	03:43	
Control Test	0.080	03:43	
Air Blank	0.000	03:43	
Subject Sample	# 1	03:44	Sequence Aborted
Breath Volume	0.000	03:44	
Air Blank	0.000	03:44	

Data Download Date/Time: 07/11/2017 19:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DEJESUS,V
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:21

SUBJECT'S LAST NAME: SAYRE

DATE/TIME OF TEST: 06/07/2017 03:52:10

Diagnostics Check	OK	03:45
Air Blank	0.000	03:46
Control Test	0.080	03:46
Air Blank	0.000	03:46
Subject Sample # 1	0.077	03:47
Breath Volume	1.851	03:47
Air Blank	0.000	03:48
Air Blank	0.000	03:50
Subject Sample # 2	0.079	03:50
Breath Volume	2.898	03:50
Air Blank	0.000	03:51
Control Test	0.080	03:51
Air Blank	0.000	03:52
Diagnostics Check	OK	03:52

Data Download Date/Time: 07/11/2017 19:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARLAUG, G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:37

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 06/08/2017 04:03:54

Diagnostics Check	OK	03:57
Air Blank	0.000	03:58
Control Test	0.080	03:58
Air Blank	0.000	03:58
Subject Sample # 1	0.187	03:59
Breath Volume	1.585	03:59
Air Blank	0.000	04:00
Air Blank	0.000	04:01
Subject Sample # 2	0.189	04:02
Breath Volume	2.406	04:02
Air Blank	0.000	04:02
Control Test	0.080	04:03
Air Blank	0.000	04:03
Diagnostics Check	OK	04:03

Data Download Date/Time: 07/11/2017 19:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK,CASEY,S
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARLAUG,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: VALDES

DATE/TIME OF TEST: 06/09/2017 03:17:15

Diagnostics Check	OK	03:10
Air Blank	0.000	03:11
Control Test	0.081	03:11
Air Blank	0.000	03:12
Subject Sample # 1	0.232	03:12
Breath Volume	1.816	03:12
Air Blank	0.000	03:13
Air Blank	0.000	03:15
Subject Sample # 2	0.230	03:15
Breath Volume	1.191	03:15
Air Blank	0.000	03:16
Control Test	0.080	03:16
Air Blank	0.000	03:17
Diagnostics Check	OK	03:17

Data Download Date/Time: 07/11/2017 19:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: BENITEZ TORRES

DATE/TIME OF TEST: 06/09/2017 04:38:09

Diagnostics Check	OK	04:30
Air Blank	0.000	04:31
Control Test	0.081	04:31
Air Blank	0.000	04:32
Subject Sample # 1	0.126	04:33
Breath Volume	2.386	04:33
Air Blank	0.000	04:34
Air Blank	0.000	04:35
Subject Sample # 2	0.118	04:36
Breath Volume	1.351	04:36
Air Blank	0.000	04:37
Control Test	0.080	04:37
Air Blank	0.000	04:38
Diagnostics Check	OK	04:38

Data Download Date/Time: 07/11/2017 19:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK,CASEY,S
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:58

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 06/10/2017 00:24:44

Diagnostics Check	OK	00:18
Air Blank	0.000	00:18
Control Test	0.080	00:19
Air Blank	0.000	00:19
Subject Sample # 1	0.172	00:20
Breath Volume	1.558	00:20
Air Blank	0.000	00:20
Air Blank	0.000	00:22
Subject Sample # 2	0.159	00:23
Breath Volume	1.242	00:23
Air Blank	0.000	00:23
Control Test	0.080	00:24
Air Blank	0.000	00:24
Diagnostics Check	OK	00:24

Data Download Date/Time: 07/11/2017 19:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRIMBLE,TERESA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BLANCHARD,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:44

SUBJECT'S LAST NAME: KENNEY

DATE/TIME OF TEST: 06/11/2017 03:13:21

Diagnostics Check	OK	03:04	
Air Blank	0.000	03:05	
Control Test	0.081	03:05	
Air Blank	0.000	03:05	
Subject Sample	# 1	03:09	No Sample Provided
Breath Volume	0.000	03:09	
Air Blank	0.000	03:09	
Air Blank	0.000	03:11	
Subject Sample	# 2	03:11	Subject Test Refused
Breath Volume	0.000	03:11	
Air Blank	0.000	03:12	
Control Test	0.081	03:12	
Air Blank	0.000	03:13	
Diagnostics Check	OK	03:13	

Data Download Date/Time: 07/11/2017 19:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: TAVERAS, N
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: SANTIAGO MADERA

DATE/TIME OF TEST: 06/11/2017 03:32:19

Diagnostics Check	OK	03:26
Air Blank	0.000	03:26
Control Test	0.081	03:26
Air Blank	0.000	03:27
Subject Sample # 1	0.212	03:27
Breath Volume	1.527	03:27
Air Blank	0.000	03:28
Air Blank	0.000	03:30
Subject Sample # 2	0.210	03:30
Breath Volume	1.503	03:30
Air Blank	0.000	03:31
Control Test	0.080	03:31
Air Blank	0.000	03:32
Diagnostics Check	OK	03:32

Data Download Date/Time: 07/11/2017 19:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GONZALEZ,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:02

SUBJECT'S LAST NAME: JUAREZ

DATE/TIME OF TEST: 06/11/2017 05:29:15

Diagnostics Check	OK	05:22
Air Blank	0.000	05:22
Control Test	0.080	05:23
Air Blank	0.000	05:23
Subject Sample # 1	0.146	05:24
Breath Volume	1.945	05:24
Air Blank	0.000	05:25
Air Blank	0.000	05:27
Subject Sample # 2	0.153	05:27
Breath Volume	2.136	05:27
Air Blank	0.000	05:28
Control Test	0.079	05:28
Air Blank	0.000	05:29
Diagnostics Check	OK	05:29

Data Download Date/Time: 07/11/2017 19:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNIDER,P
ARRESTING OFFICER AGENCY: Plant City PD
LAST AGENCY INSPECTION: 06/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:49

SUBJECT'S LAST NAME: GUTIERREZ-DUENAS

DATE/TIME OF TEST: 06/12/2017 02:19:26

Diagnostics Check	OK	02:09
Air Blank	0.000	02:10
Control Test	0.080	02:10
Air Blank	0.000	02:10
Subject Sample # 1	0.113	02:14
Breath Volume	1.175	02:14
Air Blank	0.000	02:14
Air Blank	0.000	02:16
Subject Sample # 2	0.127	02:17
Breath Volume	1.968	02:17
Air Blank	0.000	02:18
Control Test	0.078	02:18
Air Blank	0.000	02:19
Diagnostics Check	OK	02:19

Data Download Date/Time: 07/11/2017 19:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VELAZQUEZ,B
ARRESTING OFFICER AGENCY: U of So FL Campus PD
LAST AGENCY INSPECTION: 06/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 06/13/2017 02:26:35

Diagnostics Check	OK	02:20
Air Blank	0.000	02:20
Control Test	0.080	02:21
Air Blank	0.000	02:21
Subject Sample # 1	0.179	02:22
Breath Volume	2.585	02:22
Air Blank	0.000	02:22
Air Blank	0.000	02:24
Subject Sample # 2	0.173	02:25
Breath Volume	2.808	02:25
Air Blank	0.000	02:25
Control Test	0.079	02:25
Air Blank	0.000	02:26
Diagnostics Check	OK	02:26

Data Download Date/Time: 07/11/2017 19:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK,CASEY,S
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BLANCHARD,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11117080A3
DRY GAS EXPIRY DATE: 06/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:53

SUBJECT'S LAST NAME: AVILES JR

DATE/TIME OF TEST: 06/14/2017 01:20:06

Diagnostics Check	OK	01:14
Air Blank	0.000	01:14
Control Test	0.080	01:14
Air Blank	0.000	01:15
Subject Sample # 1	0.104	01:15
Breath Volume	3.253	01:15
Air Blank	0.000	01:16
Air Blank	0.000	01:18
Subject Sample # 2	0.101	01:18
Breath Volume	2.996	01:18
Air Blank	0.000	01:19
Control Test	0.079	01:19
Air Blank	0.000	01:20
Diagnostics Check	OK	01:20

Data Download Date/Time: 07/11/2017 19:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK,CASEY,S
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GONZALEZ,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11117080A3
DRY GAS EXPIRY DATE: 06/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 06/14/2017 23:52:00

Diagnostics Check	OK	23:45
Air Blank	0.000	23:46
Control Test	0.080	23:46
Air Blank	0.000	23:46
Subject Sample # 1	0.166	23:47
Breath Volume	3.042	23:47
Air Blank	0.000	23:48
Air Blank	0.000	23:49
Subject Sample # 2	0.161	23:50
Breath Volume	2.941	23:50
Air Blank	0.000	23:51
Control Test	0.079	23:51
Air Blank	0.000	23:51
Diagnostics Check	OK	23:52

Data Download Date/Time: 07/11/2017 19:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: NAVE,P
ARRESTING OFFICER AGENCY: Plant City PD
LAST AGENCY INSPECTION: 06/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11117080A3
DRY GAS EXPIRY DATE: 06/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:11

SUBJECT'S LAST NAME: WRIGHT

DATE/TIME OF TEST: 06/15/2017 04:39:28

Diagnostics Check	OK	04:32
Air Blank	0.000	04:32
Control Test	0.080	04:32
Air Blank	0.000	04:33
Subject Sample # 1	0.156	04:34
Breath Volume	1.839	04:34
Air Blank	0.000	04:35
Air Blank	0.000	04:36
Subject Sample # 2	0.160	04:37
Breath Volume	1.554	04:37
Air Blank	0.000	04:38
Control Test	0.078	04:38
Air Blank	0.000	04:39
Diagnostics Check	OK	04:39

Data Download Date/Time: 07/11/2017 19:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARLAUG, G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11117080A3
DRY GAS EXPIRY DATE: 06/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:37

SUBJECT'S LAST NAME: SCOTT

DATE/TIME OF TEST: 06/16/2017 00:04:37

Diagnostics Check	OK	23:58
Air Blank	0.000	23:59
Control Test	0.079	23:59
Air Blank	0.000	23:59
Subject Sample # 1	0.178	00:00
Breath Volume	1.308	00:00
Air Blank	0.000	00:00
Air Blank	0.000	00:02
Subject Sample # 2	0.190	00:03
Breath Volume	1.609	00:03
Air Blank	0.000	00:03
Control Test	0.079	00:04
Air Blank	0.000	00:04
Diagnostics Check	OK	00:04

Data Download Date/Time: 07/11/2017 19:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GONZALEZ,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11117080A3
DRY GAS EXPIRY DATE: 06/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:43

SUBJECT'S LAST NAME: KERR

DATE/TIME OF TEST: 06/16/2017 01:06:52

Diagnostics Check	OK	01:03	
Air Blank	0.000	01:04	
Control Test	0.079	01:04	
Air Blank	0.000	01:04	
Subject Sample	# 1	01:05	Subject Test Refused
Breath Volume	0.000	01:05	
Air Blank	0.000	01:05	
Control Test	0.080	01:06	
Air Blank	0.000	01:06	
Diagnostics Check	OK	01:06	

Data Download Date/Time: 07/11/2017 19:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRONSON, R
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11117080A3
DRY GAS EXPIRY DATE: 06/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:43

SUBJECT'S LAST NAME: VAZQUEZ

DATE/TIME OF TEST: 06/16/2017 23:11:48

Diagnostics Check	OK	23:03
Air Blank	0.000	23:04
Control Test	0.080	23:04
Air Blank	0.000	23:05
Subject Sample # 1	0.155	23:06
Breath Volume	2.292	23:06
Air Blank	0.000	23:07
Air Blank	0.000	23:09
Subject Sample # 2	0.138	23:10
Breath Volume	1.875	23:10
Air Blank	0.000	23:10
Control Test	0.079	23:11
Air Blank	0.000	23:11
Diagnostics Check	OK	23:11

Data Download Date/Time: 07/11/2017 19:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRIMBLE,TERESA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11117080A3
DRY GAS EXPIRY DATE: 06/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:31

SUBJECT'S LAST NAME: BOLDEN

DATE/TIME OF TEST: 06/16/2017 23:57:47

Diagnostics Check	OK	23:51
Air Blank	0.000	23:51
Control Test	0.080	23:52
Air Blank	0.000	23:52
Subject Sample # 1	0.073	23:53
Breath Volume	2.316	23:53
Air Blank	0.000	23:53
Air Blank	0.000	23:55
Subject Sample # 2	0.073	23:56
Breath Volume	1.664	23:56
Air Blank	0.000	23:56
Control Test	0.079	23:57
Air Blank	0.000	23:57
Diagnostics Check	OK	23:57

Data Download Date/Time: 07/11/2017 19:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRIMBLE,TERESA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: CRAGG,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11117080A3
DRY GAS EXPIRY DATE: 06/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:06

SUBJECT'S LAST NAME: HOCKETT

DATE/TIME OF TEST: 06/17/2017 03:38:36

Diagnostics Check	OK	03:26	
Air Blank	0.000	03:27	
Control Test	0.079	03:27	
Air Blank	0.000	03:28	
Subject Sample	# 1 0.103	03:31	Volume Not Met
Breath Volume	0.152	03:31	
Air Blank	0.000	03:32	
Air Blank	0.000	03:33	
Subject Sample	# 2 0.126	03:36	Volume Not Met
Breath Volume	0.015	03:36	
Air Blank	0.000	03:37	
Control Test	0.077	03:37	
Air Blank	0.000	03:38	
Diagnostics Check	OK	03:38	

Data Download Date/Time: 07/11/2017 19:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK,CASEY,S
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: CRAGG,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11117080A3
DRY GAS EXPIRY DATE: 06/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:36

SUBJECT'S LAST NAME: LLANES

DATE/TIME OF TEST: 06/17/2017 23:12:20

Diagnostics Check	OK	23:00	
Air Blank	0.000	23:01	
Control Test	0.079	23:01	
Air Blank	0.000	23:01	
Subject Sample	# 1 0.334	23:05	Volume Not Met
Breath Volume	0.566	23:05	
Air Blank	0.000	23:05	
Air Blank	0.000	23:07	
Subject Sample	# 2 0.310	23:10	Volume Not Met
Breath Volume	0.253	23:10	
Air Blank	0.000	23:11	
Control Test	0.076	23:11	
Air Blank	0.000	23:12	
Diagnostics Check	OK	23:12	

Data Download Date/Time: 07/11/2017 19:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SIMPKINS,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11117080A3
DRY GAS EXPIRY DATE: 06/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:28

SUBJECT'S LAST NAME: GOMEZ-GOMEZ

DATE/TIME OF TEST: 06/18/2017 03:02:20

Diagnostics Check	OK	02:48	
Air Blank	0.000	02:49	
Control Test	0.079	02:49	
Air Blank	0.000	02:49	
Subject Sample	# 1 0.250	02:52	Volume Not Met
Breath Volume	0.015	02:52	
Air Blank	0.000	02:53	
Air Blank	0.000	02:55	
Subject Sample	# 2 0.265	02:57	
Breath Volume	1.773	02:57	
Air Blank	0.000	02:58	
Air Blank	0.000	03:00	
Subject Sample	# 3 0.258	03:00	
Breath Volume	1.265	03:00	
Air Blank	0.000	03:01	
Control Test	0.078	03:01	
Air Blank	0.000	03:02	
Diagnostics Check	OK	03:02	

Data Download Date/Time: 07/11/2017 19:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GONZALEZ,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11117080A3
DRY GAS EXPIRY DATE: 06/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:51

SUBJECT'S LAST NAME: WALKER

DATE/TIME OF TEST: 06/18/2017 04:19:57

Diagnostics Check	OK	04:12
Air Blank	0.000	04:12
Control Test	0.078	04:12
Air Blank	0.000	04:13
Subject Sample # 1	0.254	04:14
Breath Volume	1.128	04:14
Air Blank	0.000	04:14
Air Blank	0.000	04:16
Subject Sample # 2	0.261	04:18
Breath Volume	1.109	04:18
Air Blank	0.000	04:18
Control Test	0.078	04:19
Air Blank	0.000	04:19
Diagnostics Check	OK	04:19

Data Download Date/Time: 07/11/2017 19:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK,CASEY,S
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11117080A3
DRY GAS EXPIRY DATE: 06/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:54

SUBJECT'S LAST NAME: RANKIN

DATE/TIME OF TEST: 06/19/2017 02:21:34

Diagnostics Check	OK	02:15
Air Blank	0.000	02:15
Control Test	0.080	02:15
Air Blank	0.000	02:16
Subject Sample # 1	0.198	02:16
Breath Volume	2.574	02:16
Air Blank	0.000	02:17
Air Blank	0.000	02:19
Subject Sample # 2	0.187	02:19
Breath Volume	1.613	02:19
Air Blank	0.000	02:20
Control Test	0.079	02:20
Air Blank	0.000	02:21
Diagnostics Check	OK	02:21

Data Download Date/Time: 07/11/2017 19:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK,CASEY,S
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MACIAS,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11117080A3
DRY GAS EXPIRY DATE: 06/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: STEVENSON

DATE/TIME OF TEST: 06/20/2017 01:03:29

Diagnostics Check	OK	00:57
Air Blank	0.000	00:57
Control Test	0.080	00:57
Air Blank	0.000	00:58
Subject Sample # 1	0.065	00:59
Breath Volume	1.183	00:59
Air Blank	0.000	00:59
Air Blank	0.000	01:01
Subject Sample # 2	0.064	01:01
Breath Volume	1.312	01:01
Air Blank	0.000	01:02
Control Test	0.078	01:02
Air Blank	0.000	01:03
Diagnostics Check	OK	01:03

Data Download Date/Time: 07/11/2017 19:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11117080A3
DRY GAS EXPIRY DATE: 06/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:56

SUBJECT'S LAST NAME: HARDEN JR

DATE/TIME OF TEST: 06/20/2017 04:29:29

Diagnostics Check	OK	04:16	
Air Blank	0.000	04:16	
Control Test	0.080	04:16	
Air Blank	0.000	04:17	
Subject Sample	# 1 0.104	04:19	
Breath Volume	1.660	04:19	
Air Blank	0.000	04:20	
Air Blank	0.000	04:22	
Subject Sample	# 2	04:25	No Sample Provided
Breath Volume	0.000	04:25	
Air Blank	0.000	04:25	
Air Blank	0.000	04:27	
Subject Sample	# 3	04:28	Subject Test Refused
Breath Volume	0.000	04:28	
Air Blank	0.000	04:28	
Control Test	0.079	04:28	
Air Blank	0.000	04:29	
Diagnostics Check	OK	04:29	

Data Download Date/Time: 07/11/2017 19:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRONSON,R
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11117080A3
DRY GAS EXPIRY DATE: 06/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:41

SUBJECT'S LAST NAME: CUNNINGHAM

DATE/TIME OF TEST: 06/20/2017 05:07:50

Diagnostics Check	OK	05:01
Air Blank	0.000	05:02
Control Test	0.078	05:02
Air Blank	0.000	05:02
Subject Sample # 1	0.000	05:03
Breath Volume	3.679	05:03
Air Blank	0.000	05:03
Air Blank	0.000	05:05
Subject Sample # 2	0.000	05:06
Breath Volume	3.843	05:06
Air Blank	0.000	05:06
Control Test	0.080	05:07
Air Blank	0.000	05:07
Diagnostics Check	OK	05:07

Data Download Date/Time: 07/11/2017 19:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DAWSY,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11117080A3
DRY GAS EXPIRY DATE: 06/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:08

SUBJECT'S LAST NAME: KRAUSS

DATE/TIME OF TEST: 06/21/2017 05:34:09

Diagnostics Check	OK	05:28
Air Blank	0.000	05:28
Control Test	0.080	05:28
Air Blank	0.000	05:29
Subject Sample # 1	0.173	05:29
Breath Volume	2.195	05:29
Air Blank	0.000	05:30
Air Blank	0.000	05:32
Subject Sample # 2	0.175	05:32
Breath Volume	1.570	05:32
Air Blank	0.000	05:33
Control Test	0.079	05:33
Air Blank	0.000	05:34
Diagnostics Check	OK	05:34

Data Download Date/Time: 07/11/2017 19:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER, A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11117080A3
DRY GAS EXPIRY DATE: 06/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:21

SUBJECT'S LAST NAME: BLAKE

DATE/TIME OF TEST: 06/21/2017 23:48:03

Diagnostics Check	OK	23:41
Air Blank	0.000	23:42
Control Test	0.081	23:42
Air Blank	0.000	23:42
Subject Sample # 1	0.162	23:43
Breath Volume	3.195	23:43
Air Blank	0.000	23:44
Air Blank	0.000	23:46
Subject Sample # 2	0.143	23:46
Breath Volume	1.140	23:46
Air Blank	0.000	23:47
Control Test	0.080	23:47
Air Blank	0.000	23:47
Diagnostics Check	OK	23:48

Data Download Date/Time: 07/11/2017 19:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK,CASEY,S
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRONSON,R
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11117080A3
DRY GAS EXPIRY DATE: 06/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:02

SUBJECT'S LAST NAME: WIRT JR

DATE/TIME OF TEST: 06/23/2017 02:32:29

Diagnostics Check	OK	02:26
Air Blank	0.000	02:26
Control Test	0.080	02:27
Air Blank	0.000	02:27
Subject Sample # 1	0.000	02:28
Breath Volume	2.328	02:28
Air Blank	0.000	02:28
Air Blank	0.000	02:30
Subject Sample # 2	0.000	02:31
Breath Volume	1.640	02:31
Air Blank	0.000	02:31
Control Test	0.079	02:31
Air Blank	0.000	02:32
Diagnostics Check	OK	02:32

Data Download Date/Time: 07/11/2017 19:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: COSTA,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11117080A3
DRY GAS EXPIRY DATE: 06/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:00

SUBJECT'S LAST NAME: RODRIGUEZ LOPEZ

DATE/TIME OF TEST: 06/23/2017 05:29:45

Diagnostics Check	OK	05:21
Air Blank	0.000	05:21
Control Test	0.080	05:22
Air Blank	0.000	05:22
Subject Sample # 1	0.159	05:25
Breath Volume	2.437	05:25
Air Blank	0.000	05:25
Air Blank	0.000	05:27
Subject Sample # 2	0.155	05:28
Breath Volume	3.191	05:28
Air Blank	0.000	05:28
Control Test	0.077	05:29
Air Blank	0.000	05:29
Diagnostics Check	OK	05:29

Data Download Date/Time: 07/11/2017 19:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK,CASEY,S
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNIDER,P
ARRESTING OFFICER AGENCY: Plant City PD
LAST AGENCY INSPECTION: 06/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11117080A3
DRY GAS EXPIRY DATE: 06/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:31

SUBJECT'S LAST NAME: PENA MARTINEZ

DATE/TIME OF TEST: 06/24/2017 04:02:06

Diagnostics Check	OK	03:53
Air Blank	0.000	03:53
Control Test	0.080	03:54
Air Blank	0.000	03:54
Subject Sample # 1	0.154	03:57
Breath Volume	1.484	03:57
Air Blank	0.000	03:58
Air Blank	0.000	04:00
Subject Sample # 2	0.165	04:00
Breath Volume	1.363	04:00
Air Blank	0.000	04:01
Control Test	0.077	04:01
Air Blank	0.000	04:02
Diagnostics Check	OK	04:02

Data Download Date/Time: 07/11/2017 19:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARLAUG,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11117080A3
DRY GAS EXPIRY DATE: 06/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:57

SUBJECT'S LAST NAME: MALDONADO DIAZ

DATE/TIME OF TEST: 06/24/2017 04:25:17

Diagnostics Check	OK	04:17
Air Blank	0.000	04:18
Control Test	0.078	04:18
Air Blank	0.000	04:19
Subject Sample # 1	0.204	04:20
Breath Volume	1.882	04:20
Air Blank	0.000	04:21
Air Blank	0.000	04:23
Subject Sample # 2	0.199	04:23
Breath Volume	1.191	04:23
Air Blank	0.000	04:24
Control Test	0.077	04:24
Air Blank	0.000	04:25
Diagnostics Check	OK	04:25

Data Download Date/Time: 07/11/2017 19:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FRANKLIN, R
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11117080A3
DRY GAS EXPIRY DATE: 06/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:34

SUBJECT'S LAST NAME: GHELDOLF

DATE/TIME OF TEST: 06/25/2017 00:06:39

Diagnostics Check	OK	23:56
Air Blank	0.000	23:56
Control Test	0.080	23:57
Air Blank	0.000	23:57
Subject Sample # 3	0.162	23:58
Breath Volume	2.003	23:58
Air Blank	0.000	23:59
Air Blank	0.000	00:01
Subject Sample # 1	0.140	00:01
Breath Volume	1.781	00:01
Air Blank	0.000	00:02
Air Blank	0.000	00:04
Subject Sample # 2	0.152	00:05
Breath Volume	2.523	00:05
Air Blank	0.000	00:05
Control Test	0.080	00:06
Air Blank	0.000	00:06
Diagnostics Check	OK	00:06

Data Download Date/Time: 07/11/2017 19:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRABAND,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11117080A3
DRY GAS EXPIRY DATE: 06/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:03

SUBJECT'S LAST NAME: WHITE JR

DATE/TIME OF TEST: 06/25/2017 02:31:10

Diagnostics Check	OK	02:24
Air Blank	0.000	02:25
Control Test	0.080	02:25
Air Blank	0.000	02:26
Subject Sample # 1	0.075	02:26
Breath Volume	3.992	02:26
Air Blank	0.000	02:27
Air Blank	0.000	02:29
Subject Sample # 2	0.079	02:29
Breath Volume	3.511	02:29
Air Blank	0.000	02:30
Control Test	0.079	02:30
Air Blank	0.000	02:31
Diagnostics Check	OK	02:31

Data Download Date/Time: 07/11/2017 19:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: NIEUWENDAM, R
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11117080A3
DRY GAS EXPIRY DATE: 06/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: GOMEZ

DATE/TIME OF TEST: 06/25/2017 02:51:54

Diagnostics Check	OK	02:45
Air Blank	0.000	02:45
Control Test	0.079	02:45
Air Blank	0.000	02:46
Subject Sample # 1	0.105	02:47
Breath Volume	2.328	02:47
Air Blank	0.000	02:47
Air Blank	0.000	02:49
Subject Sample # 2	0.101	02:50
Breath Volume	1.542	02:50
Air Blank	0.000	02:50
Control Test	0.079	02:51
Air Blank	0.000	02:51
Diagnostics Check	OK	02:51

Data Download Date/Time: 07/11/2017 19:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRIMBLE,TERESA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GONZALEZ,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11117080A3
DRY GAS EXPIRY DATE: 06/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:27

SUBJECT'S LAST NAME: CASAS GALLEGOS

DATE/TIME OF TEST: 06/25/2017 03:53:51

Diagnostics Check	OK	03:47
Air Blank	0.000	03:47
Control Test	0.079	03:48
Air Blank	0.000	03:48
Subject Sample # 1	0.192	03:49
Breath Volume	2.105	03:49
Air Blank	0.000	03:49
Air Blank	0.000	03:51
Subject Sample # 2	0.194	03:52
Breath Volume	3.058	03:52
Air Blank	0.000	03:52
Control Test	0.079	03:53
Air Blank	0.000	03:53
Diagnostics Check	OK	03:53

Data Download Date/Time: 07/11/2017 19:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VAN TREESE,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11117080A3
DRY GAS EXPIRY DATE: 06/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:21

SUBJECT'S LAST NAME: CARDONA GUERRERO

DATE/TIME OF TEST: 06/25/2017 22:48:02

Diagnostics Check	OK	22:41
Air Blank	0.000	22:41
Control Test	0.080	22:42
Air Blank	0.000	22:42
Subject Sample # 1	0.140	22:43
Breath Volume	1.390	22:43
Air Blank	0.000	22:43
Air Blank	0.000	22:45
Subject Sample # 2	0.142	22:46
Breath Volume	1.707	22:46
Air Blank	0.000	22:47
Control Test	0.079	22:47
Air Blank	0.000	22:47
Diagnostics Check	OK	22:48

Data Download Date/Time: 07/11/2017 19:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FRANKLIN, R
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11117080A3
DRY GAS EXPIRY DATE: 06/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:42

SUBJECT'S LAST NAME: PRICE

DATE/TIME OF TEST: 06/25/2017 23:09:30

Diagnostics Check	OK	23:02
Air Blank	0.000	23:02
Control Test	0.080	23:03
Air Blank	0.000	23:03
Subject Sample # 1	0.211	23:04
Breath Volume	1.539	23:04
Air Blank	0.000	23:05
Air Blank	0.000	23:07
Subject Sample # 2	0.192	23:07
Breath Volume	2.121	23:07
Air Blank	0.000	23:08
Control Test	0.078	23:08
Air Blank	0.000	23:09
Diagnostics Check	OK	23:09

Data Download Date/Time: 07/11/2017 19:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK,CASEY,S
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: COSTA,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11117080A3
DRY GAS EXPIRY DATE: 06/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: DE LA ROSA

DATE/TIME OF TEST: 06/28/2017 02:30:13

Diagnostics Check	OK	02:22
Air Blank	0.000	02:23
Control Test	0.080	02:23
Air Blank	0.000	02:23
Subject Sample # 1	0.097	02:25
Breath Volume	1.367	02:25
Air Blank	0.000	02:26
Air Blank	0.000	02:27
Subject Sample # 2	0.097	02:28
Breath Volume	1.414	02:28
Air Blank	0.000	02:29
Control Test	0.078	02:29
Air Blank	0.000	02:30
Diagnostics Check	OK	02:30

Data Download Date/Time: 07/11/2017 19:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BLANCHARD, J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11117080A3
DRY GAS EXPIRY DATE: 06/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:22

SUBJECT'S LAST NAME: DANIELS JR

DATE/TIME OF TEST: 06/30/2017 02:54:59

Diagnostics Check	OK	02:43	
Air Blank	0.000	02:43	
Control Test	0.080	02:44	
Air Blank	0.000	02:44	
Subject Sample	# 1 0.196	02:47	Volume Not Met
Breath Volume	0.312	02:47	
Air Blank	0.000	02:48	
Air Blank	0.000	02:50	
Subject Sample	# 2 0.197	02:53	Volume Not Met
Breath Volume	0.933	02:53	
Air Blank	0.000	02:54	
Control Test	0.079	02:54	
Air Blank	0.000	02:54	
Diagnostics Check	OK	02:54	

Data Download Date/Time: 07/11/2017 19:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003406

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KELLER,DAVID,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: KELLER,DAVID
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 06/07/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:28

SUBJECT'S LAST NAME: LAMB

DATE/TIME OF TEST: 06/10/2017 03:05:03

Diagnostics Check	OK	02:58
Air Blank	0.000	02:59
Control Test	0.081	02:59
Air Blank	0.000	03:00
Subject Sample # 1	0.193	03:00
Breath Volume	1.882	03:00
Air Blank	0.000	03:01
Air Blank	0.000	03:03
Subject Sample # 2	0.185	03:03
Breath Volume	1.855	03:03
Air Blank	0.000	03:04
Control Test	0.081	03:04
Air Blank	0.000	03:04
Diagnostics Check	OK	03:05

Data Download Date/Time: 07/13/2017 10:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003406

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KELLER,DAVID,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: KELLER,DAVID
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 06/07/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:28

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 06/11/2017 02:56:59

Diagnostics Check	OK	02:51
Air Blank	0.000	02:51
Control Test	0.081	02:51
Air Blank	0.000	02:52
Subject Sample # 1	0.194	02:52
Breath Volume	2.125	02:52
Air Blank	0.000	02:53
Air Blank	0.000	02:55
Subject Sample # 2	0.186	02:55
Breath Volume	3.164	02:55
Air Blank	0.000	02:56
Control Test	0.081	02:56
Air Blank	0.000	02:56
Diagnostics Check	OK	02:56

Data Download Date/Time: 07/13/2017 10:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003406

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEAS,BRANDYN,M
BREATH TEST OPERATOR AGENCY: Leon County SO
ARRESTING OFFICER: DEAS,BRANDYN
ARRESTING OFFICER AGENCY: Leon County SO
LAST AGENCY INSPECTION: 06/07/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: HARRELL

DATE/TIME OF TEST: 06/11/2017 04:32:31

Diagnostics Check	OK	04:26
Air Blank	0.000	04:26
Control Test	0.081	04:26
Air Blank	0.000	04:27
Subject Sample # 1	0.119	04:27
Breath Volume	2.050	04:27
Air Blank	0.000	04:28
Air Blank	0.000	04:30
Subject Sample # 2	0.118	04:31
Breath Volume	1.937	04:31
Air Blank	0.000	04:31
Control Test	0.081	04:31
Air Blank	0.000	04:32
Diagnostics Check	OK	04:32

Data Download Date/Time: 07/13/2017 10:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003406

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILSON,BRANDON,J
BREATH TEST OPERATOR AGENCY: FL State University
ARRESTING OFFICER: WILSON,BRANDON
ARRESTING OFFICER AGENCY: FL State University
LAST AGENCY INSPECTION: 06/07/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: KERWIN

DATE/TIME OF TEST: 06/25/2017 03:05:57

Diagnostics Check	OK	02:59
Air Blank	0.000	02:59
Control Test	0.081	03:00
Air Blank	0.000	03:00
Subject Sample # 1	0.087	03:01
Breath Volume	1.437	03:01
Air Blank	0.000	03:02
Air Blank	0.000	03:03
Subject Sample # 2	0.091	03:04
Breath Volume	1.300	03:04
Air Blank	0.000	03:05
Control Test	0.081	03:05
Air Blank	0.000	03:05
Diagnostics Check	OK	03:05

Data Download Date/Time: 07/13/2017 10:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003406

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVERETT, MATTHEW, A
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: NORTHWAY, G
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 06/07/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 05:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:57

SUBJECT'S LAST NAME: SERRANO-INIGUEZ

DATE/TIME OF TEST: 06/25/2017 06:26:53

Diagnostics Check	OK	06:20
Air Blank	0.000	06:20
Control Test	0.081	06:20
Air Blank	0.000	06:21
Subject Sample # 1	0.242	06:22
Breath Volume	1.371	06:22
Air Blank	0.000	06:22
Air Blank	0.000	06:24
Subject Sample # 2	0.250	06:25
Breath Volume	2.253	06:25
Air Blank	0.000	06:26
Control Test	0.080	06:26
Air Blank	0.000	06:26
Diagnostics Check	OK	06:26

Data Download Date/Time: 07/13/2017 10:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003406

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROY,G,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: ROY,G
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 06/07/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:20

SUBJECT'S LAST NAME: HENLEY

DATE/TIME OF TEST: 06/28/2017 20:49:15

Diagnostics Check	OK	20:42
Air Blank	0.000	20:43
Control Test	0.081	20:43
Air Blank	0.000	20:44
Subject Sample # 1	0.191	20:44
Breath Volume	1.910	20:44
Air Blank	0.000	20:45
Air Blank	0.000	20:47
Subject Sample # 2	0.205	20:47
Breath Volume	1.824	20:47
Air Blank	0.000	20:48
Control Test	0.081	20:48
Air Blank	0.000	20:49
Diagnostics Check	OK	20:49

Data Download Date/Time: 07/13/2017 10:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003407

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,DANIEL,M
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: MARTIN,DANIEL
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 05/27/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:44

SUBJECT'S LAST NAME: STANSEL

DATE/TIME OF TEST: 06/07/2017 00:31:41

Diagnostics Check	OK	00:28	
Air Blank	0.000	00:29	
Control Test	0.080	00:29	
Air Blank	0.000	00:30	
Subject Sample	# 1	00:30	Subject Test Refused
Breath Volume	0.000	00:30	
Air Blank	0.000	00:30	
Control Test	0.079	00:31	
Air Blank	0.000	00:31	
Diagnostics Check	OK	00:31	

Data Download Date/Time: 06/29/2017 00:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003408

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANCHEZ,MARGARITA,
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: DEPESTRE,GREGORY
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 01:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: MANUEL

DATE/TIME OF TEST: 06/04/2017 04:04:40

Diagnostics Check	OK	03:51	
Air Blank	0.000	03:51	
Control Test	0.076	03:51	
Air Blank	0.000	03:52	
Subject Sample	# 1 0.122	03:55	Volume Not Met
Breath Volume	0.304	03:55	
Air Blank	0.000	03:56	
Air Blank	0.000	03:57	
Subject Sample	# 2 0.126	03:58	
Breath Volume	1.429	03:58	
Air Blank	0.000	03:59	
Air Blank	0.000	04:01	
Subject Sample	# 3 0.125	04:03	
Breath Volume	1.613	04:03	
Air Blank	0.000	04:03	
Control Test	0.075	04:04	
Air Blank	0.000	04:04	
Diagnostics Check	OK	04:04	

Data Download Date/Time: 07/07/2017 02:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003408

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,ARGENIS,A
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: PEREZ,ARGENIS
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 06:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:15

SUBJECT'S LAST NAME: VICENTE VICENTE

DATE/TIME OF TEST: 06/04/2017 07:05:17

Diagnostics Check	OK	06:58
Air Blank	0.000	06:59
Control Test	0.077	06:59
Air Blank	0.000	07:00
Subject Sample # 1	0.127	07:00
Breath Volume	1.847	07:00
Air Blank	0.000	07:01
Air Blank	0.000	07:03
Subject Sample # 2	0.122	07:03
Breath Volume	2.292	07:03
Air Blank	0.000	07:04
Control Test	0.077	07:04
Air Blank	0.000	07:05
Diagnostics Check	OK	07:05

Data Download Date/Time: 07/07/2017 02:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003408

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASTILLO,JOSUET,
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: CASTILLO,JOSUET
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 08:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:17

SUBJECT'S LAST NAME: VERA

DATE/TIME OF TEST: 06/08/2017 09:24:14

Diagnostics Check	OK	09:09	
Air Blank	0.000	09:10	
Control Test	0.080	09:10	
Air Blank	0.000	09:10	
Subject Sample	# 1 0.125	09:14	Volume Not Met
Breath Volume	0.804	09:14	
Air Blank	0.000	09:14	
Air Blank	0.000	09:16	
Subject Sample	# 2 0.143	09:17	
Breath Volume	1.519	09:17	
Air Blank	0.000	09:17	
Air Blank	0.000	09:19	
Subject Sample	# 3 0.125	09:22	Volume Not Met
Breath Volume	0.050	09:22	
Air Blank	0.000	09:23	
Control Test	0.079	09:23	
Air Blank	0.000	09:24	
Diagnostics Check	OK	09:24	

Data Download Date/Time: 07/07/2017 02:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003408

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASTILLO,JOSUET,
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: CASTILLO,JOSUET
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 08:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:17

SUBJECT'S LAST NAME: VERA

DATE/TIME OF TEST: 06/08/2017 09:35:38

Diagnostics Check	OK	09:26
Air Blank	0.000	09:26
Control Test	0.078	09:26
Air Blank	0.000	09:27
Subject Sample # 1	0.141	09:29
Breath Volume	1.343	09:29
Air Blank	0.000	09:29
Air Blank	0.000	09:31
Subject Sample # 2	0.141	09:34
Breath Volume	1.167	09:34
Air Blank	0.000	09:34
Control Test	0.078	09:35
Air Blank	0.000	09:35
Diagnostics Check	OK	09:35

Data Download Date/Time: 07/07/2017 02:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003408

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANTANA LOPEZ,ERNESTO,
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: SANTANA LOPEZ,ERNESTO
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 03:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: RAMOS

DATE/TIME OF TEST: 06/10/2017 04:03:12

Diagnostics Check	OK	03:57
Air Blank	0.000	03:57
Control Test	0.080	03:57
Air Blank	0.000	03:58
Subject Sample # 1	0.161	03:58
Breath Volume	2.773	03:58
Air Blank	0.000	03:59
Air Blank	0.000	04:01
Subject Sample # 2	0.168	04:01
Breath Volume	2.820	04:01
Air Blank	0.000	04:02
Control Test	0.079	04:02
Air Blank	0.000	04:03
Diagnostics Check	OK	04:03

Data Download Date/Time: 07/07/2017 02:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003408

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VIVEIROS,KYLE,M
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: GOMEZ,ANTONIO
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 02:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:44

SUBJECT'S LAST NAME: ZAMORA

DATE/TIME OF TEST: 06/10/2017 04:21:43

Diagnostics Check	OK	04:11
Air Blank	0.000	04:11
Control Test	0.080	04:12
Air Blank	0.000	04:12
Subject Sample # 1	0.143	04:15
Breath Volume	1.324	04:15
Air Blank	0.000	04:16
Air Blank	0.000	04:18
Subject Sample # 2	0.141	04:20
Breath Volume	1.570	04:20
Air Blank	0.000	04:20
Control Test	0.079	04:21
Air Blank	0.000	04:21
Diagnostics Check	OK	04:21

Data Download Date/Time: 07/07/2017 02:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003408

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HICKS,TERENCE,E
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: HICKS,TERENCE
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 06:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:05

SUBJECT'S LAST NAME: RESTREPO

DATE/TIME OF TEST: 06/17/2017 08:32:07

Diagnostics Check	OK	08:29	
Air Blank	0.000	08:29	
Control Test	0.081	08:30	
Air Blank	0.000	08:30	
Subject Sample	# 1	08:30	Subject Test Refused
Breath Volume	0.000	08:30	
Air Blank	0.000	08:31	
Control Test	0.080	08:31	
Air Blank	0.000	08:32	
Diagnostics Check	OK	08:32	

Data Download Date/Time: 07/07/2017 02:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003408

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LASSO SILVA,NICOLAS,
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: LASSO SILVA,NICOLAS
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 19:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:09

SUBJECT'S LAST NAME: GISBERT

DATE/TIME OF TEST: 06/18/2017 19:48:04

Diagnostics Check	OK	19:41
Air Blank	0.000	19:42
Control Test	0.081	19:42
Air Blank	0.000	19:43
Subject Sample # 1	0.223	19:43
Breath Volume	1.753	19:43
Air Blank	0.000	19:44
Air Blank	0.000	19:46
Subject Sample # 2	0.213	19:46
Breath Volume	1.753	19:46
Air Blank	0.000	19:47
Control Test	0.081	19:47
Air Blank	0.000	19:47
Diagnostics Check	OK	19:48

Data Download Date/Time: 07/07/2017 02:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003408

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASTILLO,JOSUET,
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: CASTILLO,JOSUET
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 01:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 06/21/2017 02:50:10

Diagnostics Check	OK	02:41	
Air Blank	0.000	02:42	
Control Test	0.082	02:42	
Air Blank	0.000	02:42	
Subject Sample	# 1	02:46	No Sample Provided
Breath Volume	0.000	02:46	
Air Blank	0.000	02:46	
Air Blank	0.000	02:48	
Subject Sample	# 2	02:48	Subject Test Refused
Breath Volume	0.000	02:48	
Air Blank	0.000	02:49	
Control Test	0.081	02:49	
Air Blank	0.000	02:50	
Diagnostics Check	OK	02:50	

Data Download Date/Time: 07/07/2017 02:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003408

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASTILLO,JOSUET,
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: CASTILLO,JOSUET
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 02:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: GONZALEZ PONCIANO

DATE/TIME OF TEST: 06/22/2017 03:53:51

Diagnostics Check	OK	03:46
Air Blank	0.000	03:47
Control Test	0.082	03:47
Air Blank	0.000	03:48
Subject Sample # 1	0.135	03:49
Breath Volume	2.527	03:49
Air Blank	0.000	03:49
Air Blank	0.000	03:51
Subject Sample # 2	0.137	03:52
Breath Volume	1.367	03:52
Air Blank	0.000	03:52
Control Test	0.082	03:53
Air Blank	0.000	03:53
Diagnostics Check	OK	03:53

Data Download Date/Time: 07/07/2017 02:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003408

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASTILLO,JOSUET,
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: CASTILLO,JOSUET
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 03:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:12

SUBJECT'S LAST NAME: SALDANA

DATE/TIME OF TEST: 06/24/2017 04:40:33

Diagnostics Check	OK	04:33
Air Blank	0.000	04:33
Control Test	0.082	04:34
Air Blank	0.000	04:34
Subject Sample # 1	0.104	04:36
Breath Volume	1.941	04:36
Air Blank	0.000	04:36
Air Blank	0.000	04:38
Subject Sample # 2	0.100	04:39
Breath Volume	1.476	04:39
Air Blank	0.000	04:39
Control Test	0.081	04:39
Air Blank	0.000	04:40
Diagnostics Check	OK	04:40

Data Download Date/Time: 07/07/2017 02:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003408

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIVERO,JOSE,I
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: RIVERO,JOSE
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 00:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:47

SUBJECT'S LAST NAME: VIERA

DATE/TIME OF TEST: 06/25/2017 01:21:26

Diagnostics Check	OK	01:14
Air Blank	0.000	01:15
Control Test	0.081	01:15
Air Blank	0.000	01:16
Subject Sample # 1	0.259	01:16
Breath Volume	2.417	01:16
Air Blank	0.000	01:17
Air Blank	0.000	01:19
Subject Sample # 2	0.251	01:19
Breath Volume	2.500	01:19
Air Blank	0.000	01:20
Control Test	0.080	01:20
Air Blank	0.000	01:21
Diagnostics Check	OK	01:21

Data Download Date/Time: 07/07/2017 02:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003408

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRIEGO, CHARLENE, R
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: LLANIO, PIERRE
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 07:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:50

SUBJECT'S LAST NAME: GONZALEZ PINO

DATE/TIME OF TEST: 06/25/2017 08:24:20

Diagnostics Check	OK	08:18
Air Blank	0.000	08:18
Control Test	0.081	08:19
Air Blank	0.000	08:19
Subject Sample # 1	0.244	08:19
Breath Volume	1.804	08:19
Air Blank	0.000	08:20
Air Blank	0.000	08:22
Subject Sample # 2	0.244	08:22
Breath Volume	2.308	08:22
Air Blank	0.000	08:23
Control Test	0.080	08:23
Air Blank	0.000	08:24
Diagnostics Check	OK	08:24

Data Download Date/Time: 07/07/2017 02:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003408

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRIEGO, CHARLENE, R
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: GOMEZ, ANTONIO
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 07:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:20

SUBJECT'S LAST NAME: MENDEZ PABLO

DATE/TIME OF TEST: 06/26/2017 08:47:49

Diagnostics Check	OK	08:41
Air Blank	0.000	08:41
Control Test	0.080	08:42
Air Blank	0.000	08:42
Subject Sample # 1	0.184	08:43
Breath Volume	3.617	08:43
Air Blank	0.000	08:44
Air Blank	0.000	08:45
Subject Sample # 2	0.179	08:46
Breath Volume	1.542	08:46
Air Blank	0.000	08:46
Control Test	0.080	08:47
Air Blank	0.000	08:47
Diagnostics Check	OK	08:47

Data Download Date/Time: 07/07/2017 02:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003413

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLERBEE,JOHN,E
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: ELLERBEE,JOHN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 06/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A3
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:15

SUBJECT'S LAST NAME: BARNES

DATE/TIME OF TEST: 06/12/2017 20:56:44

Diagnostics Check	OK	20:50
Air Blank	0.000	20:51
Control Test	0.079	20:51
Air Blank	0.000	20:52
Subject Sample # 1	0.069	20:52
Breath Volume	2.421	20:52
Air Blank	0.000	20:53
Air Blank	0.000	20:54
Subject Sample # 2	0.064	20:55
Breath Volume	2.917	20:55
Air Blank	0.000	20:55
Control Test	0.078	20:56
Air Blank	0.000	20:56
Diagnostics Check	OK	20:56

Data Download Date/Time: 07/28/2017 11:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003937

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BEEKMAN,JACOB,C
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: BEEKMAN,JACOB
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 05/24/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 03:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: DEJESUS-LOPEZ

DATE/TIME OF TEST: 06/14/2017 04:47:06

Diagnostics Check	OK	04:40
Air Blank	0.000	04:41
Control Test	0.080	04:41
Air Blank	0.000	04:41
Subject Sample # 1	0.161	04:42
Breath Volume	2.710	04:42
Air Blank	0.000	04:43
Air Blank	0.000	04:45
Subject Sample # 2	0.160	04:45
Breath Volume	2.660	04:45
Air Blank	0.000	04:46
Control Test	0.081	04:46
Air Blank	0.000	04:47
Diagnostics Check	OK	04:47

Data Download Date/Time: 06/27/2017 20:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003937

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: STINSON,ANDREW,L

BREATH TEST OPERATOR AGENCY: Kissimmee PD

ARRESTING OFFICER: STINSON,ANDREW

ARRESTING OFFICER AGENCY: Kissimmee PD

LAST AGENCY INSPECTION: 05/24/2017

Date or Time Changed: No

DRY GAS STANDARD LOT #: 163437

DRY GAS EXPIRY DATE: 10/30/2018

VIOLATION CODE: DUI

ARREST TIME: 01:51

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: GAINES JR

DATE/TIME OF TEST: 06/16/2017 02:27:43

Diagnostics Check	OK	02:21
Air Blank	0.000	02:21
Control Test	0.080	02:21
Air Blank	0.000	02:22
Subject Sample # 1	0.113	02:23
Breath Volume	2.515	02:23
Air Blank	0.000	02:23
Air Blank	0.000	02:25
Subject Sample # 2	0.108	02:26
Breath Volume	2.785	02:26
Air Blank	0.000	02:26
Control Test	0.079	02:27
Air Blank	0.000	02:27
Diagnostics Check	OK	02:27

Data Download Date/Time: 06/27/2017 20:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004587

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BEEKMAN,JACOB,C
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: BEEKMAN,JACOB
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 05/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 22:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:08

SUBJECT'S LAST NAME: CLARK

DATE/TIME OF TEST: 06/08/2017 22:43:54

Diagnostics Check	OK	22:36	
Air Blank	0.000	22:37	
Control Test	0.078	22:37	
Air Blank	0.000	22:38	
Subject Sample	# 1 0.000	22:39	
Breath Volume	3.242	22:39	
Air Blank	0.000	22:40	
Air Blank	0.000	22:42	
Subject Sample	# 2	22:42	Slope Not Met
Breath Volume	2.730	22:42	
Air Blank	0.000	22:43	
Control Test	0.077	22:43	
Air Blank	0.000	22:43	
Diagnostics Check	OK	22:43	

Data Download Date/Time: 06/27/2017 21:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004587

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BEEKMAN,JACOB,C
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: BEEKMAN,JACOB
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 05/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 22:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:08

SUBJECT'S LAST NAME: CLARK

DATE/TIME OF TEST: 06/08/2017 22:54:44

Diagnostics Check	OK	22:48	
Air Blank	0.000	22:48	
Control Test	0.077	22:48	
Air Blank	0.000	22:49	
Subject Sample # 1	0.000	22:50	
Breath Volume	1.871	22:50	
Air Blank	0.000	22:50	
Air Blank	0.000	22:52	
Subject Sample # 2	0.000	22:53	
Breath Volume	2.179	22:53	
Air Blank	0.000	22:53	
Control Test	0.030	22:54	Interferent Detect
Air Blank	0.000	22:54	

Data Download Date/Time: 06/27/2017 21:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004587

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BEEKMAN,JACOB,C
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: BEEKMAN,JACOB
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 05/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 03:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: DEJESUS-LOPEZ

DATE/TIME OF TEST: 06/14/2017 03:31:42

Diagnostics Check	OK	03:29	
Air Blank	0.000	03:30	
Control Test	0.077	03:30	
Air Blank		03:31	Ambient Fail
Air Blank		03:31	Ambient Fail

Data Download Date/Time: 06/27/2017 21:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004587

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BEEKMAN,JACOB,C
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: BEEKMAN,JACOB
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 05/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 03:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: DEJESUS-LOPEZ

DATE/TIME OF TEST: 06/14/2017 03:40:27

Diagnostics Check	OK	03:33	
Air Blank	0.000	03:34	
Control Test	0.077	03:34	
Air Blank	0.000	03:35	
Subject Sample	# 1 0.179	03:35	
Breath Volume	2.554	03:35	
Air Blank	0.000	03:36	
Air Blank	0.000	03:38	
Subject Sample	# 2 0.177	03:38	
Breath Volume	2.960	03:38	
Air Blank	0.000	03:39	
Control Test	0.074	03:39	Control Outside Tolerance
Air Blank	0.000	03:40	

Data Download Date/Time: 06/27/2017 21:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004587

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPOS,MICHAEL,A
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: CAMPOS,MICHAEL
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 06/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 23:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:03

SUBJECT'S LAST NAME: HANCOCK

DATE/TIME OF TEST: 06/30/2017 23:47:27

Diagnostics Check	OK	23:41
Air Blank	0.000	23:41
Control Test	0.078	23:41
Air Blank	0.000	23:42
Subject Sample # 1	0.121	23:43
Breath Volume	1.832	23:43
Air Blank	0.000	23:43
Air Blank	0.000	23:45
Subject Sample # 2	0.114	23:45
Breath Volume	1.421	23:45
Air Blank	0.000	23:46
Control Test	0.078	23:46
Air Blank	0.000	23:47
Diagnostics Check	OK	23:47

Data Download Date/Time: 07/11/2017 15:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,WILLIE,G
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: ROGRIGUEZ,PAUL
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 05/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 01:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:36

SUBJECT'S LAST NAME: ROJAS

DATE/TIME OF TEST: 06/04/2017 01:59:47

Diagnostics Check	OK	01:56	
Air Blank	0.000	01:57	
Control Test	0.081	01:57	
Air Blank	0.000	01:58	
Subject Sample	# 1	01:58	Subject Test Refused
Breath Volume	0.000	01:58	
Air Blank	0.000	01:58	
Control Test	0.081	01:59	
Air Blank	0.000	01:59	
Diagnostics Check	OK	01:59	

Data Download Date/Time: 06/29/2017 12:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SARMENT,JOSEPH,D
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: SARMENT,JOSEPH
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 05/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 20:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:47

SUBJECT'S LAST NAME: OLSON

DATE/TIME OF TEST: 06/05/2017 22:22:01

Diagnostics Check	OK	22:14
Air Blank	0.000	22:14
Control Test	0.082	22:15
Air Blank	0.000	22:15
Subject Sample # 1	0.103	22:17
Breath Volume	2.007	22:17
Air Blank	0.000	22:18
Air Blank	0.000	22:19
Subject Sample # 2	0.105	22:20
Breath Volume	2.105	22:20
Air Blank	0.000	22:21
Control Test	0.080	22:21
Air Blank	0.000	22:21
Diagnostics Check	OK	22:22

Data Download Date/Time: 06/29/2017 12:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOGGAN,MARLON,M
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: KRISER,SHANE
ARRESTING OFFICER AGENCY: Crestview PD
LAST AGENCY INSPECTION: 05/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 20:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: PORTUGAL

DATE/TIME OF TEST: 06/11/2017 22:22:30

Diagnostics Check	OK	22:11	
Air Blank	0.000	22:11	
Control Test	0.082	22:11	
Air Blank	0.000	22:12	
Subject Sample	# 1	22:15	No Sample Provided
Breath Volume	0.000	22:15	
Air Blank	0.000	22:15	
Air Blank	0.000	22:18	
Subject Sample	# 2	22:21	No Sample Provided
Breath Volume	0.000	22:21	
Air Blank	0.000	22:21	
Control Test	0.082	22:21	
Air Blank	0.000	22:22	
Diagnostics Check	OK	22:22	

Data Download Date/Time: 06/29/2017 12:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SARMENT,JOSEPH,D
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: SARMENT,JOSEPH
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 05/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 23:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:17

SUBJECT'S LAST NAME: MILANO

DATE/TIME OF TEST: 06/12/2017 00:50:55

Diagnostics Check	OK	00:44	
Air Blank	0.000	00:44	
Control Test	0.081	00:44	
Air Blank	0.000	00:45	
Subject Sample	# 1 0.226	00:46	
Breath Volume	2.847	00:46	
Air Blank	0.000	00:47	
Air Blank	0.000	00:49	
Subject Sample	# 2	00:49	Subject Test Refused
Breath Volume	0.000	00:49	
Air Blank	0.000	00:50	
Control Test	0.079	00:50	
Air Blank	0.000	00:50	
Diagnostics Check	OK	00:50	

Data Download Date/Time: 06/29/2017 12:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRUNDIN,NICHOLAS,O
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: CLARK,CHARLES
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 05/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: RICE

DATE/TIME OF TEST: 06/14/2017 02:57:27

Diagnostics Check	OK	02:51
Air Blank	0.000	02:51
Control Test	0.080	02:52
Air Blank	0.000	02:52
Subject Sample # 1	0.000	02:53
Breath Volume	2.660	02:53
Air Blank	0.000	02:53
Air Blank	0.000	02:55
Subject Sample # 2	0.000	02:56
Breath Volume	3.972	02:56
Air Blank	0.000	02:56
Control Test	0.080	02:56
Air Blank	0.000	02:57
Diagnostics Check	OK	02:57

Data Download Date/Time: 06/29/2017 12:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADKINS,JOSEPH,L
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: ADKINS,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 05/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 12:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: HILL

DATE/TIME OF TEST: 06/19/2017 01:30:01

Diagnostics Check	OK	01:24
Air Blank	0.000	01:24
Control Test	0.081	01:24
Air Blank	0.000	01:25
Subject Sample # 1	0.164	01:25
Breath Volume	2.128	01:25
Air Blank	0.000	01:26
Air Blank	0.000	01:28
Subject Sample # 2	0.161	01:28
Breath Volume	2.292	01:28
Air Blank	0.000	01:29
Control Test	0.081	01:29
Air Blank	0.000	01:29
Diagnostics Check	OK	01:30

Data Download Date/Time: 06/29/2017 12:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005076

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FOWLER,TYLER,J
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: FOWLER,TYLER
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 05/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:48

SUBJECT'S LAST NAME: CLEMONS

DATE/TIME OF TEST: 06/02/2017 02:16:56

Diagnostics Check	OK	02:11	
Air Blank	0.000	02:11	
Control Test	0.078	02:12	
Air Blank	0.000	02:12	
Subject Sample	# 1 0.138	02:13	
Breath Volume	2.468	02:13	
Air Blank	0.000	02:13	
Air Blank	0.000	02:15	
Subject Sample	# 2 0.134	02:15	
Breath Volume	2.515	02:15	
Air Blank	0.000	02:16	
Control Test	0.000	02:16	RFI Detect
Air Blank		02:16	RFI Detect

Data Download Date/Time: 06/28/2017 10:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005076

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FOWLER, TYLER, J
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: FOWLER, TYLER
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 05/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:48

SUBJECT'S LAST NAME: CLEMONS

DATE/TIME OF TEST: 06/02/2017 02:24:41

Diagnostics Check	OK	02:18
Air Blank	0.000	02:19
Control Test	0.078	02:19
Air Blank	0.000	02:19
Subject Sample # 1	0.129	02:20
Breath Volume	2.324	02:20
Air Blank	0.000	02:21
Air Blank	0.000	02:22
Subject Sample # 2	0.127	02:23
Breath Volume	2.167	02:23
Air Blank	0.000	02:23
Control Test	0.077	02:24
Air Blank	0.000	02:24
Diagnostics Check	OK	02:24

Data Download Date/Time: 06/28/2017 10:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005076

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FOWLER, TYLER, J
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: FOWLER, TYLER
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 05/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:09

SUBJECT'S LAST NAME: BRICKER

DATE/TIME OF TEST: 06/15/2017 01:37:39

Diagnostics Check	OK	01:30
Air Blank	0.000	01:31
Control Test	0.078	01:31
Air Blank	0.000	01:32
Subject Sample # 1	0.157	01:32
Breath Volume	1.496	01:32
Air Blank	0.000	01:33
Air Blank	0.000	01:35
Subject Sample # 2	0.171	01:36
Breath Volume	2.839	01:36
Air Blank	0.000	01:36
Control Test	0.076	01:37
Air Blank	0.000	01:37
Diagnostics Check	OK	01:37

Data Download Date/Time: 06/28/2017 10:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005076

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FOWLER, TYLER, J
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: FOWLER, TYLER
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 05/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: CALDERON

DATE/TIME OF TEST: 06/17/2017 22:48:57

Diagnostics Check	OK	22:41
Air Blank	0.000	22:41
Control Test	0.079	22:41
Air Blank	0.000	22:42
Subject Sample # 1	0.136	22:43
Breath Volume	1.824	22:43
Air Blank	0.000	22:44
Air Blank	0.000	22:45
Subject Sample # 2	0.141	22:47
Breath Volume	1.671	22:47
Air Blank	0.000	22:48
Control Test	0.077	22:48
Air Blank	0.000	22:48
Diagnostics Check	OK	22:48

Data Download Date/Time: 06/28/2017 10:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005076

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FOWLER, TYLER, J
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: FOWLER, TYLER
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 05/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: FOX

DATE/TIME OF TEST: 06/22/2017 22:32:55

Diagnostics Check	OK	22:25
Air Blank	0.000	22:26
Control Test	0.077	22:26
Air Blank	0.000	22:27
Subject Sample # 1	0.073	22:28
Breath Volume	2.195	22:28
Air Blank	0.000	22:28
Air Blank	0.000	22:30
Subject Sample # 2	0.076	22:31
Breath Volume	2.667	22:31
Air Blank	0.000	22:32
Control Test	0.077	22:32
Air Blank	0.000	22:32
Diagnostics Check	OK	22:32

Data Download Date/Time: 06/28/2017 10:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005076

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FOWLER, TYLER, J
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: FOWLER, TYLER
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 05/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:54

SUBJECT'S LAST NAME: ESCOBAR GONZALES

DATE/TIME OF TEST: 06/24/2017 01:24:48

Diagnostics Check	OK	01:14	
Air Blank	0.000	01:14	
Control Test	0.077	01:15	
Air Blank	0.000	01:15	
Subject Sample	# 1 0.154	01:18	Volume Not Met
Breath Volume	0.082	01:18	
Air Blank	0.000	01:19	
Air Blank	0.000	01:21	
Subject Sample	# 2	01:23	Subject Test Refused
Breath Volume	0.019	01:23	
Air Blank	0.000	01:23	
Control Test	0.078	01:24	
Air Blank	0.000	01:24	
Diagnostics Check	OK	01:24	

Data Download Date/Time: 06/28/2017 10:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MACIAS, A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:56

SUBJECT'S LAST NAME: AVALOS-MONTAMO

DATE/TIME OF TEST: 06/06/2017 23:23:17

Diagnostics Check	OK	23:16
Air Blank	0.000	23:17
Control Test	0.081	23:17
Air Blank	0.000	23:18
Subject Sample # 1	0.007	23:18
Breath Volume	1.863	23:18
Air Blank	0.000	23:19
Air Blank	0.000	23:21
Subject Sample # 2	0.000	23:21
Breath Volume	1.910	23:21
Air Blank	0.000	23:22
Control Test	0.081	23:22
Air Blank	0.000	23:23
Diagnostics Check	OK	23:23

Data Download Date/Time: 06/22/2017 11:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRIMBLE,TERESA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GONZALEZ,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: TURNER

DATE/TIME OF TEST: 06/07/2017 02:38:34

Diagnostics Check	OK	02:31
Air Blank	0.000	02:32
Control Test	0.081	02:32
Air Blank	0.000	02:33
Subject Sample # 1	0.292	02:33
Breath Volume	2.578	02:33
Air Blank	0.000	02:34
Air Blank	0.000	02:36
Subject Sample # 2	0.307	02:36
Breath Volume	2.324	02:36
Air Blank	0.000	02:37
Control Test	0.081	02:37
Air Blank	0.000	02:38
Diagnostics Check	OK	02:38

Data Download Date/Time: 06/22/2017 11:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRIMBLE,TERESA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BUCHER,P
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 18:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:14

SUBJECT'S LAST NAME: DEMITRO

DATE/TIME OF TEST: 06/07/2017 19:41:15

Diagnostics Check	OK	19:34
Air Blank	0.000	19:35
Control Test	0.081	19:35
Air Blank	0.000	19:36
Subject Sample # 1	0.020	19:36
Breath Volume	2.511	19:36
Air Blank	0.000	19:37
Air Blank	0.000	19:39
Subject Sample # 2	0.018	19:39
Breath Volume	2.804	19:39
Air Blank	0.000	19:40
Control Test	0.081	19:40
Air Blank	0.000	19:41
Diagnostics Check	OK	19:41

Data Download Date/Time: 06/22/2017 11:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DAWSEY,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:54

SUBJECT'S LAST NAME: STUART

DATE/TIME OF TEST: 06/10/2017 01:20:40

Diagnostics Check	OK	01:14
Air Blank	0.000	01:14
Control Test	0.082	01:14
Air Blank	0.000	01:15
Subject Sample # 1	0.000	01:16
Breath Volume	1.484	01:16
Air Blank	0.000	01:16
Air Blank	0.000	01:18
Subject Sample # 2	0.000	01:19
Breath Volume	1.218	01:19
Air Blank	0.000	01:19
Control Test	0.082	01:20
Air Blank	0.000	01:20
Diagnostics Check	OK	01:20

Data Download Date/Time: 06/22/2017 11:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARLAUG,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:28

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 06/10/2017 04:56:17

Diagnostics Check	OK	04:49
Air Blank	0.000	04:49
Control Test	0.081	04:49
Air Blank	0.000	04:50
Subject Sample # 1	0.221	04:50
Breath Volume	1.617	04:50
Air Blank	0.000	04:51
Air Blank	0.000	04:53
Subject Sample # 2	0.224	04:54
Breath Volume	1.765	04:54
Air Blank	0.000	04:55
Control Test	0.080	04:55
Air Blank	0.000	04:56
Diagnostics Check	OK	04:56

Data Download Date/Time: 06/22/2017 11:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARR,BARBARA,H
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MATAS,T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 13:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:09

SUBJECT'S LAST NAME: ALVAREZ

DATE/TIME OF TEST: 06/10/2017 16:37:26

Diagnostics Check	OK	16:30
Air Blank	0.000	16:31
Control Test	0.081	16:31
Air Blank	0.000	16:32
Subject Sample # 1	0.271	16:32
Breath Volume	2.746	16:32
Air Blank	0.000	16:33
Air Blank	0.000	16:35
Subject Sample # 2	0.272	16:35
Breath Volume	2.015	16:35
Air Blank	0.000	16:36
Control Test	0.081	16:36
Air Blank	0.000	16:37
Diagnostics Check	OK	16:37

Data Download Date/Time: 06/22/2017 11:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRIMBLE,TERESA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:42

SUBJECT'S LAST NAME: TRIPPLET

DATE/TIME OF TEST: 06/11/2017 00:14:45

Diagnostics Check	OK	00:02	
Air Blank	0.000	00:03	
Control Test	0.081	00:03	
Air Blank	0.000	00:04	
Subject Sample	# 1	00:07	No Sample Provided
Breath Volume	0.675	00:07	
Air Blank	0.000	00:07	
Air Blank	0.000	00:09	
Subject Sample	# 2	0.258	00:10
Breath Volume	1.730	00:10	
Air Blank	0.000	00:10	
Air Blank	0.000	00:12	
Subject Sample	# 3	0.259	00:13
Breath Volume	1.542	00:13	
Air Blank	0.000	00:13	
Control Test	0.081	00:14	
Air Blank	0.000	00:14	
Diagnostics Check	OK	00:14	

Data Download Date/Time: 06/22/2017 11:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRONSON,R
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:04

SUBJECT'S LAST NAME: RIVERA NEGRON

DATE/TIME OF TEST: 06/11/2017 03:31:56

Diagnostics Check	OK	03:24
Air Blank	0.000	03:25
Control Test	0.081	03:25
Air Blank	0.000	03:26
Subject Sample # 1	0.147	03:27
Breath Volume	1.480	03:27
Air Blank	0.000	03:28
Air Blank	0.000	03:29
Subject Sample # 2	0.160	03:30
Breath Volume	1.273	03:30
Air Blank	0.000	03:31
Control Test	0.081	03:31
Air Blank	0.000	03:31
Diagnostics Check	OK	03:31

Data Download Date/Time: 06/22/2017 11:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRIMBLE,TERESA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PORTMAN,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: CAMPFIELD

DATE/TIME OF TEST: 06/11/2017 04:02:46

Diagnostics Check	OK	03:55
Air Blank	0.000	03:56
Control Test	0.081	03:56
Air Blank	0.000	03:57
Subject Sample # 1	0.082	03:58
Breath Volume	3.937	03:58
Air Blank	0.000	03:58
Air Blank	0.000	04:00
Subject Sample # 2	0.078	04:01
Breath Volume	3.898	04:01
Air Blank	0.000	04:01
Control Test	0.082	04:02
Air Blank	0.000	04:02
Diagnostics Check	OK	04:02

Data Download Date/Time: 06/22/2017 11:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DAWSY, B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:23

SUBJECT'S LAST NAME: DUMERVIL

DATE/TIME OF TEST: 06/11/2017 04:51:26

Diagnostics Check	OK	04:43
Air Blank	0.000	04:44
Control Test	0.081	04:44
Air Blank	0.000	04:45
Subject Sample # 1	0.079	04:47
Breath Volume	1.105	04:47
Air Blank	0.000	04:47
Air Blank	0.000	04:49
Subject Sample # 2	0.077	04:49
Breath Volume	1.152	04:49
Air Blank	0.000	04:50
Control Test	0.082	04:50
Air Blank	0.000	04:51
Diagnostics Check	OK	04:51

Data Download Date/Time: 06/22/2017 11:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRIMBLE,TERESA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 05:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:06

SUBJECT'S LAST NAME: HERNANDEZ RAMIREZ

DATE/TIME OF TEST: 06/11/2017 06:33:51

Diagnostics Check	OK	06:26
Air Blank	0.000	06:27
Control Test	0.081	06:27
Air Blank	0.000	06:27
Subject Sample # 1	0.221	06:29
Breath Volume	1.980	06:29
Air Blank	0.000	06:29
Air Blank	0.000	06:31
Subject Sample # 2	0.212	06:32
Breath Volume	1.687	06:32
Air Blank	0.000	06:32
Control Test	0.081	06:33
Air Blank	0.000	06:33
Diagnostics Check	OK	06:33

Data Download Date/Time: 06/22/2017 11:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: TAVARES,N
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: AYALA VISSEPO

DATE/TIME OF TEST: 06/12/2017 04:35:35

Diagnostics Check	OK	04:30	
Air Blank	0.000	04:30	
Control Test	0.082	04:31	
Air Blank	0.000	04:31	
Subject Sample	# 1	04:34	Subject Test Refused
Breath Volume	0.300	04:34	
Air Blank	0.000	04:34	
Control Test	0.082	04:34	
Air Blank	0.000	04:35	
Diagnostics Check	OK	04:35	

Data Download Date/Time: 06/22/2017 11:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SIMPKINS, D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:42

SUBJECT'S LAST NAME: HAMIDA

DATE/TIME OF TEST: 06/13/2017 04:09:57

Diagnostics Check	OK	04:02
Air Blank	0.000	04:02
Control Test	0.081	04:03
Air Blank	0.000	04:03
Subject Sample # 1	0.123	04:04
Breath Volume	2.218	04:04
Air Blank	0.000	04:05
Air Blank	0.000	04:07
Subject Sample # 2	0.125	04:08
Breath Volume	2.570	04:08
Air Blank	0.000	04:09
Control Test	0.082	04:09
Air Blank	0.000	04:09
Diagnostics Check	OK	04:09

Data Download Date/Time: 06/22/2017 11:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARLAUG,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:32

SUBJECT'S LAST NAME: KRYSA

DATE/TIME OF TEST: 06/14/2017 20:55:42

Diagnostics Check	OK	20:52	
Air Blank	0.000	20:53	
Control Test	0.081	20:53	
Air Blank	0.000	20:54	
Subject Sample	# 1	20:54	Subject Test Refused
Breath Volume	0.000	20:54	
Air Blank	0.000	20:54	
Control Test	0.081	20:55	
Air Blank	0.000	20:55	
Diagnostics Check	OK	20:55	

Data Download Date/Time: 06/22/2017 11:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:49

SUBJECT'S LAST NAME: LOPEZ LOPEZ

DATE/TIME OF TEST: 06/15/2017 05:16:32

Diagnostics Check	OK	05:09
Air Blank	0.000	05:10
Control Test	0.081	05:10
Air Blank	0.000	05:11
Subject Sample # 1	0.215	05:12
Breath Volume	3.613	05:12
Air Blank	0.000	05:12
Air Blank	0.000	05:14
Subject Sample # 2	0.217	05:14
Breath Volume	3.425	05:14
Air Blank	0.000	05:15
Control Test	0.081	05:15
Air Blank	0.000	05:16
Diagnostics Check	OK	05:16

Data Download Date/Time: 06/22/2017 11:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKIPPER,ROGER,G
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: NA,N
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: ADMIN
ARREST TIME: 11:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:40

SUBJECT'S LAST NAME: ATENZA

DATE/TIME OF TEST: 06/15/2017 15:22:31

Diagnostics Check	OK	15:19
Air Blank	0.000	15:19
Control Test	0.081	15:20
Air Blank	0.000	15:20
Subject Sample # 1	0.369	15:21
Breath Volume	1.199	15:21
Air Blank	0.000	15:21
Air Blank	0.000	15:22

Data Download Date/Time: 06/22/2017 11:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRIMBLE,TERESA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GRACE,K
ARRESTING OFFICER AGENCY: Seminole Res (Hills)
LAST AGENCY INSPECTION: 06/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: IANNUCELLI

DATE/TIME OF TEST: 06/15/2017 22:38:32

Diagnostics Check	OK	22:31
Air Blank	0.000	22:31
Control Test	0.081	22:31
Air Blank	0.000	22:32
Subject Sample # 1	0.188	22:33
Breath Volume	2.871	22:33
Air Blank	0.000	22:34
Air Blank	0.000	22:36
Subject Sample # 2	0.184	22:36
Breath Volume	2.359	22:36
Air Blank	0.000	22:37
Control Test	0.080	22:37
Air Blank	0.000	22:38
Diagnostics Check	OK	22:38

Data Download Date/Time: 06/22/2017 11:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRIMBLE,TERESA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRENNAN,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:44

SUBJECT'S LAST NAME: VAIL

DATE/TIME OF TEST: 06/16/2017 00:09:29

Diagnostics Check	OK	00:04	
Air Blank	0.000	00:05	
Control Test	0.081	00:05	
Air Blank	0.000	00:06	
Subject Sample	# 1	00:08	Subject Test Refused
Breath Volume	0.000	00:08	
Air Blank	0.000	00:08	
Control Test	0.081	00:08	
Air Blank	0.000	00:09	
Diagnostics Check	OK	00:09	

Data Download Date/Time: 06/22/2017 11:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNIDER, P
ARRESTING OFFICER AGENCY: Plant City PD
LAST AGENCY INSPECTION: 06/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:49

SUBJECT'S LAST NAME: JIMENEZ

DATE/TIME OF TEST: 06/16/2017 01:17:13

Diagnostics Check	OK	01:09
Air Blank	0.000	01:10
Control Test	0.081	01:10
Air Blank	0.000	01:10
Subject Sample # 1	0.180	01:12
Breath Volume	3.929	01:12
Air Blank	0.000	01:13
Air Blank	0.000	01:14
Subject Sample # 2	0.186	01:15
Breath Volume	1.718	01:15
Air Blank	0.000	01:16
Control Test	0.081	01:16
Air Blank	0.000	01:17
Diagnostics Check	OK	01:17

Data Download Date/Time: 06/22/2017 11:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FRANKLIN,R
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:27

SUBJECT'S LAST NAME: CARMONA

DATE/TIME OF TEST: 06/16/2017 02:54:14

Diagnostics Check	OK	02:47
Air Blank	0.000	02:47
Control Test	0.082	02:48
Air Blank	0.000	02:48
Subject Sample # 1	0.189	02:49
Breath Volume	1.593	02:49
Air Blank	0.000	02:50
Air Blank	0.000	02:52
Subject Sample # 2	0.192	02:52
Breath Volume	1.742	02:52
Air Blank	0.000	02:53
Control Test	0.080	02:53
Air Blank	0.000	02:54
Diagnostics Check	OK	02:54

Data Download Date/Time: 06/22/2017 11:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PALFY, R
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:31

SUBJECT'S LAST NAME: BERGMANN

DATE/TIME OF TEST: 06/16/2017 23:57:51

Diagnostics Check	OK	23:51
Air Blank	0.000	23:52
Control Test	0.082	23:52
Air Blank	0.000	23:53
Subject Sample # 1	0.142	23:53
Breath Volume	1.371	23:53
Air Blank	0.000	23:54
Air Blank	0.000	23:55
Subject Sample # 2	0.128	23:56
Breath Volume	1.222	23:56
Air Blank	0.000	23:56
Control Test	0.081	23:57
Air Blank	0.000	23:57
Diagnostics Check	OK	23:57

Data Download Date/Time: 06/22/2017 11:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRIMBLE,TERESA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRENNAN,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:33

SUBJECT'S LAST NAME: BARGERON

DATE/TIME OF TEST: 06/17/2017 02:59:54

Diagnostics Check	OK	02:53
Air Blank	0.000	02:53
Control Test	0.081	02:54
Air Blank	0.000	02:54
Subject Sample # 1	0.136	02:55
Breath Volume	5.121	02:55
Air Blank	0.000	02:55
Air Blank	0.000	02:57
Subject Sample # 2	0.137	02:58
Breath Volume	5.144	02:58
Air Blank	0.000	02:58
Control Test	0.082	02:59
Air Blank	0.000	02:59
Diagnostics Check	OK	02:59

Data Download Date/Time: 06/22/2017 11:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARFIELD-CRAIG,C
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:00

SUBJECT'S LAST NAME: SNYDER

DATE/TIME OF TEST: 06/17/2017 05:27:17

Diagnostics Check	OK	05:20
Air Blank	0.000	05:21
Control Test	0.082	05:21
Air Blank	0.000	05:21
Subject Sample # 1	0.166	05:22
Breath Volume	2.699	05:22
Air Blank	0.000	05:23
Air Blank	0.000	05:25
Subject Sample # 2	0.166	05:25
Breath Volume	2.664	05:25
Air Blank	0.000	05:26
Control Test	0.082	05:26
Air Blank	0.000	05:27
Diagnostics Check	OK	05:27

Data Download Date/Time: 06/22/2017 11:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER, A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 05:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:45

SUBJECT'S LAST NAME: PINO PEREZ

DATE/TIME OF TEST: 06/17/2017 06:13:21

Diagnostics Check	OK	06:05
Air Blank	0.000	06:05
Control Test	0.081	06:06
Air Blank	0.000	06:06
Subject Sample # 1	0.144	06:08
Breath Volume	1.531	06:08
Air Blank	0.000	06:08
Air Blank	0.000	06:10
Subject Sample # 2	0.147	06:11
Breath Volume	1.867	06:11
Air Blank	0.000	06:12
Control Test	0.082	06:12
Air Blank	0.000	06:13
Diagnostics Check	OK	06:13

Data Download Date/Time: 06/22/2017 11:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: COSTA,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: TURNER

DATE/TIME OF TEST: 06/17/2017 22:11:57

Diagnostics Check	OK	22:05
Air Blank	0.000	22:06
Control Test	0.081	22:06
Air Blank	0.000	22:07
Subject Sample # 1	0.000	22:07
Breath Volume	2.332	22:07
Air Blank	0.000	22:08
Air Blank	0.000	22:10
Subject Sample # 2	0.000	22:10
Breath Volume	2.464	22:10
Air Blank	0.000	22:11
Control Test	0.082	22:11
Air Blank	0.000	22:11
Diagnostics Check	OK	22:11

Data Download Date/Time: 06/22/2017 11:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:26

SUBJECT'S LAST NAME: FEBUS SANTIAGO

DATE/TIME OF TEST: 06/18/2017 00:53:14

Diagnostics Check	OK	00:46
Air Blank	0.000	00:47
Control Test	0.082	00:47
Air Blank	0.000	00:48
Subject Sample # 1	0.156	00:48
Breath Volume	2.695	00:48
Air Blank	0.000	00:49
Air Blank	0.000	00:51
Subject Sample # 2	0.154	00:51
Breath Volume	2.308	00:51
Air Blank	0.000	00:52
Control Test	0.082	00:52
Air Blank	0.000	00:53
Diagnostics Check	OK	00:53

Data Download Date/Time: 06/22/2017 11:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DUNCAN,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:41

SUBJECT'S LAST NAME: JAMES

DATE/TIME OF TEST: 06/18/2017 02:07:41

Diagnostics Check	OK	02:01
Air Blank	0.000	02:01
Control Test	0.082	02:02
Air Blank	0.000	02:02
Subject Sample # 1	0.178	02:03
Breath Volume	2.367	02:03
Air Blank	0.000	02:03
Air Blank	0.000	02:05
Subject Sample # 2	0.185	02:06
Breath Volume	2.214	02:06
Air Blank	0.000	02:06
Control Test	0.081	02:07
Air Blank	0.000	02:07
Diagnostics Check	OK	02:07

Data Download Date/Time: 06/22/2017 11:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNYDER,MELONIE,E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SHERMAN,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 07:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:26

SUBJECT'S LAST NAME: ISLAM

DATE/TIME OF TEST: 06/18/2017 07:54:33

Diagnostics Check	OK	07:47
Air Blank	0.000	07:48
Control Test	0.082	07:48
Air Blank	0.000	07:48
Subject Sample # 1	0.032	07:49
Breath Volume	2.117	07:49
Air Blank	0.000	07:50
Air Blank	0.000	07:52
Subject Sample # 2	0.032	07:53
Breath Volume	1.960	07:53
Air Blank	0.000	07:53
Control Test	0.081	07:53
Air Blank	0.000	07:54
Diagnostics Check	OK	07:54

Data Download Date/Time: 06/22/2017 11:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,REBECCA,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: UNO,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 07:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:42

SUBJECT'S LAST NAME: SANCHEZ

DATE/TIME OF TEST: 06/18/2017 08:11:16

Diagnostics Check	OK	08:04
Air Blank	0.000	08:05
Control Test	0.081	08:05
Air Blank	0.000	08:06
Subject Sample # 1	0.201	08:06
Breath Volume	2.304	08:06
Air Blank	0.000	08:07
Air Blank	0.000	08:09
Subject Sample # 2	0.196	08:09
Breath Volume	2.500	08:09
Air Blank	0.000	08:10
Control Test	0.080	08:10
Air Blank	0.000	08:11
Diagnostics Check	OK	08:11

Data Download Date/Time: 06/22/2017 11:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: COSTA,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:37

SUBJECT'S LAST NAME: STINE

DATE/TIME OF TEST: 06/19/2017 01:03:50

Diagnostics Check	OK	00:57
Air Blank	0.000	00:57
Control Test	0.081	00:57
Air Blank	0.000	00:58
Subject Sample # 1	0.218	00:59
Breath Volume	3.355	00:59
Air Blank	0.000	00:59
Air Blank	0.000	01:01
Subject Sample # 2	0.227	01:02
Breath Volume	2.925	01:02
Air Blank	0.000	01:02
Control Test	0.081	01:03
Air Blank	0.000	01:03
Diagnostics Check	OK	01:03

Data Download Date/Time: 06/22/2017 11:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PAULINO,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:28

SUBJECT'S LAST NAME: SPRINGSTEAD

DATE/TIME OF TEST: 06/19/2017 03:58:36

Diagnostics Check	OK	03:48	
Air Blank	0.000	03:48	
Control Test	0.082	03:48	
Air Blank	0.000	03:49	
Subject Sample	# 1	03:52	No Sample Provided
Breath Volume	0.000	03:52	
Air Blank	0.000	03:53	
Air Blank	0.000	03:55	
Subject Sample	# 2	03:57	Subject Test Refused
Breath Volume	0.000	03:57	
Air Blank	0.000	03:57	
Control Test	0.082	03:58	
Air Blank	0.000	03:58	
Diagnostics Check	OK	03:58	

Data Download Date/Time: 06/22/2017 11:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRONSON,R
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:43

SUBJECT'S LAST NAME: DUDLEY

DATE/TIME OF TEST: 06/19/2017 20:10:17

Diagnostics Check	OK	20:04
Air Blank	0.000	20:04
Control Test	0.081	20:04
Air Blank	0.000	20:05
Subject Sample # 1	0.261	20:05
Breath Volume	2.781	20:05
Air Blank	0.000	20:06
Air Blank	0.000	20:08
Subject Sample # 2	0.255	20:08
Breath Volume	1.812	20:08
Air Blank	0.000	20:09
Control Test	0.080	20:09
Air Blank	0.000	20:10
Diagnostics Check	OK	20:10

Data Download Date/Time: 06/22/2017 11:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRENNAN, D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 18:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:17

SUBJECT'S LAST NAME: GUTHRIE

DATE/TIME OF TEST: 06/20/2017 19:50:47

Diagnostics Check	OK	19:38	
Air Blank	0.000	19:39	
Control Test	0.082	19:39	
Air Blank	0.000	19:40	
Subject Sample	# 1 0.159	19:43	Volume Not Met
Breath Volume	0.386	19:43	
Air Blank	0.000	19:44	
Air Blank	0.000	19:45	
Subject Sample	# 2 0.184	19:48	Volume Not Met
Breath Volume	0.117	19:48	
Air Blank	0.000	19:49	
Control Test	0.083	19:50	
Air Blank	0.000	19:50	
Diagnostics Check	OK	19:50	

Data Download Date/Time: 06/22/2017 11:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRIMBLE,TERESA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: LIZARDO

DATE/TIME OF TEST: 06/21/2017 04:47:39

Diagnostics Check	OK	04:41
Air Blank	0.000	04:41
Control Test	0.081	04:42
Air Blank	0.000	04:42
Subject Sample # 1	0.131	04:43
Breath Volume	2.597	04:43
Air Blank	0.000	04:43
Air Blank	0.000	04:45
Subject Sample # 2	0.124	04:46
Breath Volume	1.222	04:46
Air Blank	0.000	04:46
Control Test	0.081	04:47
Air Blank	0.000	04:47
Diagnostics Check	OK	04:47

Data Download Date/Time: 06/22/2017 11:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRIMBLE, TERESA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS, S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:12

SUBJECT'S LAST NAME: NOLAN

DATE/TIME OF TEST: 06/21/2017 21:40:55

Diagnostics Check	OK	21:33
Air Blank	0.000	21:33
Control Test	0.081	21:34
Air Blank	0.000	21:34
Subject Sample # 1	0.165	21:36
Breath Volume	1.218	21:36
Air Blank	0.000	21:36
Air Blank	0.000	21:38
Subject Sample # 2	0.150	21:39
Breath Volume	1.148	21:39
Air Blank	0.000	21:40
Control Test	0.081	21:40
Air Blank	0.000	21:40
Diagnostics Check	OK	21:40

Data Download Date/Time: 06/22/2017 11:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRIMBLE,TERESA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PORTMAN,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: ZAYAS

DATE/TIME OF TEST: 06/22/2017 02:07:16

Diagnostics Check	OK	02:01
Air Blank	0.000	02:01
Control Test	0.081	02:01
Air Blank	0.000	02:02
Subject Sample # 1	0.000	02:02
Breath Volume	2.667	02:02
Air Blank	0.000	02:03
Air Blank	0.000	02:05
Subject Sample # 2	0.000	02:05
Breath Volume	2.687	02:05
Air Blank	0.000	02:06
Control Test	0.082	02:06
Air Blank	0.000	02:07
Diagnostics Check	OK	02:07

Data Download Date/Time: 06/22/2017 11:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRONSON,R
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:11

SUBJECT'S LAST NAME: GARAVUSO

DATE/TIME OF TEST: 06/24/2017 01:36:21

Diagnostics Check	OK	01:31	
Air Blank	0.000	01:32	
Control Test	0.081	01:32	
Air Blank	0.000	01:32	
Subject Sample	# 1	01:34	Subject Test Refused
Breath Volume	0.000	01:34	
Air Blank	0.000	01:35	
Control Test	0.081	01:35	
Air Blank	0.000	01:36	
Diagnostics Check	OK	01:36	

Data Download Date/Time: 07/11/2017 17:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MACIAS,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: ARAGON

DATE/TIME OF TEST: 06/24/2017 02:43:19

Diagnostics Check	OK	02:40	
Air Blank	0.000	02:40	
Control Test	0.081	02:41	
Air Blank	0.000	02:41	
Subject Sample	# 1	02:41	Subject Test Refused
Breath Volume	0.000	02:41	
Air Blank	0.000	02:42	
Control Test	0.081	02:42	
Air Blank	0.000	02:43	
Diagnostics Check	OK	02:43	

Data Download Date/Time: 07/11/2017 17:28

Subject Test Electronic Data

DATE/TIME OF TEST: 06/04/2017 21:39:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005248

INSTRUMENT REGISTERED TO: FWC

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: REYNAUD,GEORGE,

BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)

ARRESTING OFFICER: BLANCO,JUAN

ARRESTING OFFICER AGENCY: FFWCC (West Palm)

LAST AGENCY INSPECTION: 05/25/2017

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1757137

DRY GAS EXPIRY DATE: 10/15/2017

VIOLATION CODE: BUI

ARREST TIME: 22:28

GENDER: Male

OBSERVATION PERIOD BEGAN: 22:28

SUBJECT'S LAST NAME: PURON

DATE/TIME OF TEST: 06/09/2017 22:58:10

Diagnostics Check	OK	22:54	
Air Blank	0.000	22:54	
Control Test	0.080	22:55	
Air Blank	0.000	22:55	
Subject Sample	# 1	22:56	Subject Test Refused
Breath Volume	0.000	22:56	
Air Blank	0.000	22:57	
Control Test	0.080	22:57	
Air Blank	0.000	22:58	
Diagnostics Check	OK	22:58	

Data Download Date/Time: 06/28/2017 11:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005248

INSTRUMENT REGISTERED TO: FWC	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1757137
BREATH TEST OPERATOR: REYNAUD,GEORGE,	DRY GAS EXPIRY DATE: 10/15/2017
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)	VIOLATION CODE: BUI
ARRESTING OFFICER: DESPIAN,MIGUEL	ARREST TIME: 18:00
ARRESTING OFFICER AGENCY: FFWCC (West Palm)	GENDER: Male
LAST AGENCY INSPECTION: 05/25/2017	OBSERVATION PERIOD BEGAN: 18:30

SUBJECT'S LAST NAME: ALVAREZ	DATE/TIME OF TEST: 06/11/2017 18:53:25
------------------------------	--

Diagnostics Check	OK	18:50	
Air Blank	0.000	18:50	
Control Test	0.079	18:51	
Air Blank	0.000	18:51	
Subject Sample	# 1	18:52	Subject Test Refused
Breath Volume	0.000	18:52	
Air Blank	0.000	18:52	
Control Test	0.079	18:52	
Air Blank	0.000	18:53	
Diagnostics Check	OK	18:53	

Data Download Date/Time: 06/28/2017 11:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,A,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ELLIS,A
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:17

SUBJECT'S LAST NAME: DEADY

DATE/TIME OF TEST: 06/01/2017 01:44:51

Diagnostics Check	OK	01:38
Air Blank	0.000	01:38
Control Test	0.082	01:38
Air Blank	0.000	01:39
Subject Sample # 1	0.050	01:40
Breath Volume	2.207	01:40
Air Blank	0.000	01:41
Air Blank	0.000	01:42
Subject Sample # 2	0.050	01:43
Breath Volume	2.359	01:43
Air Blank	0.000	01:43
Control Test	0.081	01:44
Air Blank	0.000	01:44
Diagnostics Check	OK	01:44

Data Download Date/Time: 06/28/2017 22:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LANEY,DAMON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 03:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: LUCCKAS

DATE/TIME OF TEST: 06/01/2017 04:23:50

Diagnostics Check	OK	04:17
Air Blank	0.000	04:18
Control Test	0.081	04:18
Air Blank	0.000	04:18
Subject Sample # 1	0.146	04:19
Breath Volume	2.546	04:19
Air Blank	0.000	04:20
Air Blank	0.000	04:21
Subject Sample # 2	0.142	04:22
Breath Volume	2.191	04:22
Air Blank	0.000	04:22
Control Test	0.080	04:23
Air Blank	0.000	04:23
Diagnostics Check	OK	04:23

Data Download Date/Time: 06/28/2017 22:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOWATT,NATHAN,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MOWATT,NATHAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 15:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:03

SUBJECT'S LAST NAME: HANNAH

DATE/TIME OF TEST: 06/01/2017 16:26:29

Diagnostics Check	OK	16:23	
Air Blank	0.000	16:24	
Control Test	0.083	16:24	
Air Blank	0.000	16:24	
Subject Sample	# 1	16:25	Subject Test Refused
Breath Volume	0.000	16:25	
Air Blank	0.000	16:25	
Control Test	0.082	16:25	
Air Blank	0.000	16:26	
Diagnostics Check	OK	16:26	

Data Download Date/Time: 06/28/2017 22:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUZMAN,MARTIN,V
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GUZMAN,MARTIN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 19:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:06

SUBJECT'S LAST NAME: SHAKOOR

DATE/TIME OF TEST: 06/01/2017 20:32:31

Diagnostics Check	OK	20:26
Air Blank	0.000	20:26
Control Test	0.080	20:27
Air Blank	0.000	20:27
Subject Sample # 1	0.323	20:28
Breath Volume	2.199	20:28
Air Blank	0.000	20:28
Air Blank	0.000	20:30
Subject Sample # 2	0.305	20:30
Breath Volume	1.714	20:30
Air Blank	0.000	20:31
Control Test	0.080	20:31
Air Blank	0.000	20:32
Diagnostics Check	OK	20:32

Data Download Date/Time: 06/28/2017 22:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH, WILLIAM, T
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SMITH, WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 22:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: WELLS

DATE/TIME OF TEST: 06/01/2017 23:46:17

Diagnostics Check	OK	23:39
Air Blank	0.000	23:39
Control Test	0.081	23:39
Air Blank	0.000	23:40
Subject Sample # 1	0.320	23:41
Breath Volume	1.902	23:41
Air Blank	0.000	23:42
Air Blank	0.000	23:44
Subject Sample # 2	0.323	23:44
Breath Volume	1.859	23:44
Air Blank	0.000	23:45
Control Test	0.080	23:45
Air Blank	0.000	23:46
Diagnostics Check	OK	23:46

Data Download Date/Time: 06/28/2017 22:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LANEY,DAMON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 00:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:23

SUBJECT'S LAST NAME: SIMPSON

DATE/TIME OF TEST: 06/02/2017 00:51:04

Diagnostics Check	OK	00:44
Air Blank	0.000	00:45
Control Test	0.081	00:45
Air Blank	0.000	00:45
Subject Sample # 1	0.096	00:46
Breath Volume	1.695	00:46
Air Blank	0.000	00:47
Air Blank	0.000	00:49
Subject Sample # 2	0.093	00:49
Breath Volume	1.250	00:49
Air Blank	0.000	00:50
Control Test	0.080	00:50
Air Blank	0.000	00:50
Diagnostics Check	OK	00:51

Data Download Date/Time: 06/28/2017 22:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HANNA,J
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 02:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: ACKERMAN

DATE/TIME OF TEST: 06/02/2017 04:02:52

Diagnostics Check	OK	03:50	
Air Blank	0.000	03:50	
Control Test	0.081	03:51	
Air Blank	0.000	03:51	
Subject Sample	# 1	03:54	No Sample Provided
Breath Volume	0.000	03:54	
Air Blank	0.000	03:55	
Air Blank	0.000	03:57	
Subject Sample	# 2	0.236	03:58
Breath Volume	2.246	03:58	
Air Blank	0.000	03:59	
Air Blank	0.000	04:00	
Subject Sample	# 3	0.256	04:01
Breath Volume	1.132	04:01	
Air Blank	0.000	04:01	
Control Test	0.081	04:02	
Air Blank	0.000	04:02	
Diagnostics Check	OK	04:02	

Data Download Date/Time: 06/28/2017 22:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH, WILLIAM, T
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SMITH, WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 00:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:14

SUBJECT'S LAST NAME: GARRY

DATE/TIME OF TEST: 06/03/2017 01:38:25

Diagnostics Check	OK	01:35	
Air Blank	0.000	01:35	
Control Test	0.082	01:35	
Air Blank	0.000	01:36	
Subject Sample	# 1	01:37	Subject Test Refused
Breath Volume	0.000	01:37	
Air Blank	0.000	01:37	
Control Test	0.081	01:37	
Air Blank	0.000	01:38	
Diagnostics Check	OK	01:38	

Data Download Date/Time: 06/28/2017 22:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MIDDLETON,KRISTEN,M
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: MIDDLETON,KRISTEN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 15:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:28

SUBJECT'S LAST NAME: ADKINS

DATE/TIME OF TEST: 06/03/2017 16:55:39

Diagnostics Check	OK	16:48
Air Blank	0.000	16:48
Control Test	0.081	16:49
Air Blank	0.000	16:49
Subject Sample # 1	0.268	16:50
Breath Volume	1.246	16:50
Air Blank	0.000	16:51
Air Blank	0.000	16:53
Subject Sample # 2	0.276	16:54
Breath Volume	1.804	16:54
Air Blank	0.000	16:54
Control Test	0.078	16:55
Air Blank	0.000	16:55
Diagnostics Check	OK	16:55

Data Download Date/Time: 06/28/2017 22:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KRLIN,M
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 19:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:33

SUBJECT'S LAST NAME: GRIFFIN

DATE/TIME OF TEST: 06/03/2017 21:00:49

Diagnostics Check	OK	20:54
Air Blank	0.000	20:54
Control Test	0.082	20:54
Air Blank	0.000	20:55
Subject Sample # 1	0.242	20:56
Breath Volume	1.929	20:56
Air Blank	0.000	20:56
Air Blank	0.000	20:58
Subject Sample # 2	0.255	20:59
Breath Volume	2.312	20:59
Air Blank	0.000	20:59
Control Test	0.080	21:00
Air Blank	0.000	21:00
Diagnostics Check	OK	21:00

Data Download Date/Time: 06/28/2017 22:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TEMPLE,E
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 21:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: SHUTT

DATE/TIME OF TEST: 06/03/2017 22:38:35

Diagnostics Check	OK	22:31
Air Blank	0.000	22:31
Control Test	0.081	22:31
Air Blank	0.000	22:32
Subject Sample # 1	0.177	22:34
Breath Volume	1.492	22:34
Air Blank	0.000	22:34
Air Blank	0.000	22:36
Subject Sample # 2	0.179	22:37
Breath Volume	1.605	22:37
Air Blank	0.000	22:37
Control Test	0.080	22:37
Air Blank	0.000	22:38
Diagnostics Check	OK	22:38

Data Download Date/Time: 06/28/2017 22:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,NICOLE,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: SIEM,DIMAS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 08:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:55

SUBJECT'S LAST NAME: MURRHEE

DATE/TIME OF TEST: 06/04/2017 10:22:16

Diagnostics Check	OK	10:15
Air Blank	0.000	10:16
Control Test	0.082	10:16
Air Blank	0.000	10:17
Subject Sample # 1	0.176	10:17
Breath Volume	1.742	10:17
Air Blank	0.000	10:18
Air Blank	0.000	10:20
Subject Sample # 2	0.164	10:20
Breath Volume	1.687	10:20
Air Blank	0.000	10:21
Control Test	0.081	10:21
Air Blank	0.000	10:22
Diagnostics Check	OK	10:22

Data Download Date/Time: 06/28/2017 22:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ASHCOM, TIFFANY, J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TEMPLE, ELIZABETH
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 22:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:47

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 06/04/2017 23:10:05

Diagnostics Check	OK	23:07	
Air Blank	0.000	23:07	
Control Test	0.082	23:08	
Air Blank	0.000	23:08	
Subject Sample	# 1	23:08	Subject Test Refused
Breath Volume	0.000	23:08	
Air Blank	0.000	23:09	
Control Test	0.081	23:09	
Air Blank	0.000	23:09	
Diagnostics Check	OK	23:10	

Data Download Date/Time: 06/28/2017 22:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ASHCOM, TIFFANY, J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LILLEY, K
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 00:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: KAPLAN

DATE/TIME OF TEST: 06/05/2017 02:03:07

Diagnostics Check	OK	02:00	
Air Blank	0.000	02:00	
Control Test	0.082	02:01	
Air Blank	0.000	02:01	
Subject Sample	# 1	02:01	Subject Test Refused
Breath Volume	0.000	02:01	
Air Blank	0.000	02:02	
Control Test	0.081	02:02	
Air Blank	0.000	02:03	
Diagnostics Check	OK	02:03	

Data Download Date/Time: 06/28/2017 22:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ASHCOM, TIFFANY, J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: FICK, GORDON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 06/05/2017 04:08:08

Diagnostics Check	OK	04:05	
Air Blank	0.000	04:05	
Control Test	0.081	04:06	
Air Blank	0.000	04:06	
Subject Sample	# 1	04:06	Subject Test Refused
Breath Volume	0.000	04:06	
Air Blank	0.000	04:07	
Control Test	0.081	04:07	
Air Blank	0.000	04:08	
Diagnostics Check	OK	04:08	

Data Download Date/Time: 06/28/2017 22:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUZMAN,MARTIN,V
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GUZMAN,MARTIN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 08:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:14

SUBJECT'S LAST NAME: WITT

DATE/TIME OF TEST: 06/07/2017 09:42:01

Diagnostics Check	OK	09:35
Air Blank	0.000	09:35
Control Test	0.082	09:36
Air Blank	0.000	09:36
Subject Sample # 1	0.152	09:37
Breath Volume	2.511	09:37
Air Blank	0.000	09:38
Air Blank	0.000	09:40
Subject Sample # 2	0.149	09:40
Breath Volume	1.671	09:40
Air Blank	0.000	09:41
Control Test	0.081	09:41
Air Blank	0.000	09:41
Diagnostics Check	OK	09:42

Data Download Date/Time: 06/28/2017 22:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,NICOLE,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MUNOZ JR,OBED
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 15:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:40

SUBJECT'S LAST NAME: BARKLEY JR

DATE/TIME OF TEST: 06/07/2017 16:04:25

Diagnostics Check	OK	16:00	
Air Blank	0.000	16:00	
Control Test	0.081	16:01	
Air Blank	0.000	16:01	
Subject Sample	# 1	16:03	Subject Test Refused
Breath Volume	0.000	16:03	
Air Blank	0.000	16:03	
Control Test	0.081	16:03	
Air Blank	0.000	16:04	
Diagnostics Check	OK	16:04	

Data Download Date/Time: 06/28/2017 22:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HANNA,J
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: GOOD

DATE/TIME OF TEST: 06/10/2017 02:14:59

Diagnostics Check	OK	02:06	
Air Blank	0.000	02:07	
Control Test	0.082	02:07	
Air Blank	0.000	02:07	
Subject Sample	# 1	02:10	No Sample Provided
Breath Volume	0.000	02:10	
Air Blank	0.000	02:11	
Air Blank	0.000	02:13	
Subject Sample	# 2	02:13	Subject Test Refused
Breath Volume	0.000	02:13	
Air Blank	0.000	02:14	
Control Test	0.082	02:14	
Air Blank	0.000	02:14	
Diagnostics Check	OK	02:14	

Data Download Date/Time: 06/28/2017 22:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: NOCERA,D
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 03:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 06/10/2017 04:53:21

Diagnostics Check	OK	04:50	
Air Blank	0.000	04:51	
Control Test	0.082	04:51	
Air Blank	0.000	04:51	
Subject Sample	# 1	04:51	Subject Test Refused
Breath Volume	0.000	04:51	
Air Blank	0.000	04:52	
Control Test	0.082	04:52	
Air Blank	0.000	04:53	
Diagnostics Check	OK	04:53	

Data Download Date/Time: 06/28/2017 22:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORAN,SEAN,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MORAN,SEAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/10/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:03

SUBJECT'S LAST NAME: SINGH

DATE/TIME OF TEST: 06/11/2017 03:27:02

Diagnostics Check	OK	03:24	
Air Blank	0.000	03:24	
Control Test	0.081	03:24	
Air Blank	0.000	03:25	
Subject Sample	# 1	03:25	Subject Test Refused
Breath Volume	0.000	03:25	
Air Blank	0.000	03:26	
Control Test	0.081	03:26	
Air Blank	0.000	03:26	
Diagnostics Check	OK	03:27	

Data Download Date/Time: 06/28/2017 22:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRIFFIN,KYLE,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: FOSTER,ALEXANDER
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 01:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: CARRO

DATE/TIME OF TEST: 06/12/2017 02:00:24

Diagnostics Check	OK	01:54
Air Blank	0.000	01:54
Control Test	0.082	01:54
Air Blank	0.000	01:55
Subject Sample # 1	0.179	01:55
Breath Volume	3.085	01:55
Air Blank	0.000	01:56
Air Blank	0.000	01:58
Subject Sample # 2	0.173	01:58
Breath Volume	2.546	01:58
Air Blank	0.000	01:59
Control Test	0.081	01:59
Air Blank	0.000	02:00
Diagnostics Check	OK	02:00

Data Download Date/Time: 06/28/2017 22:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRIFFIN,KYLE,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: DONOHUE,STEVEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 02:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:04

SUBJECT'S LAST NAME: FAMOUS

DATE/TIME OF TEST: 06/12/2017 03:30:31

Diagnostics Check	OK	03:26	
Air Blank	0.000	03:27	
Control Test	0.082	03:27	
Air Blank	0.000	03:28	
Subject Sample	# 1	03:29	Subject Test Refused
Breath Volume	0.000	03:29	
Air Blank	0.000	03:29	
Control Test	0.081	03:29	
Air Blank	0.000	03:30	
Diagnostics Check	OK	03:30	

Data Download Date/Time: 06/28/2017 22:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRIFFIN,KYLE,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: RICOTTILLI,JOSHUA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 03:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:38

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 06/12/2017 04:06:37

Diagnostics Check	OK	04:00
Air Blank	0.000	04:00
Control Test	0.081	04:00
Air Blank	0.000	04:01
Subject Sample # 1	0.088	04:01
Breath Volume	2.773	04:01
Air Blank	0.000	04:02
Air Blank	0.000	04:04
Subject Sample # 2	0.082	04:05
Breath Volume	2.503	04:05
Air Blank	0.000	04:05
Control Test	0.081	04:06
Air Blank	0.000	04:06
Diagnostics Check	OK	04:06

Data Download Date/Time: 06/28/2017 22:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: RICOTTILLI,J
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 23:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: STUBBS

DATE/TIME OF TEST: 06/16/2017 00:17:32

Diagnostics Check	OK	00:10
Air Blank	0.000	00:10
Control Test	0.082	00:11
Air Blank	0.000	00:11
Subject Sample # 1	0.172	00:12
Breath Volume	1.519	00:12
Air Blank	0.000	00:13
Air Blank	0.000	00:15
Subject Sample # 2	0.167	00:16
Breath Volume	1.480	00:16
Air Blank	0.000	00:16
Control Test	0.081	00:16
Air Blank	0.000	00:17
Diagnostics Check	OK	00:17

Data Download Date/Time: 06/28/2017 22:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: CRUZ,K
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 23:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: MAUL

DATE/TIME OF TEST: 06/16/2017 01:11:23

Diagnostics Check	OK	01:03
Air Blank	0.000	01:03
Control Test	0.082	01:04
Air Blank	0.000	01:04
Subject Sample # 1	0.121	01:05
Breath Volume	1.308	01:05
Air Blank	0.000	01:06
Air Blank	0.000	01:08
Subject Sample # 2	0.127	01:09
Breath Volume	1.335	01:09
Air Blank	0.000	01:10
Control Test	0.081	01:10
Air Blank	0.000	01:11
Diagnostics Check	OK	01:11

Data Download Date/Time: 06/28/2017 22:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MCMANIS,J
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: CASTELLON

DATE/TIME OF TEST: 06/17/2017 03:33:50

Diagnostics Check	OK	03:30	
Air Blank	0.000	03:31	
Control Test	0.082	03:31	
Air Blank	0.000	03:32	
Subject Sample	# 1	03:32	Subject Test Refused
Breath Volume	0.000	03:32	
Air Blank	0.000	03:32	
Control Test	0.081	03:33	
Air Blank	0.000	03:33	
Diagnostics Check	OK	03:33	

Data Download Date/Time: 06/28/2017 22:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BRITTON,S
ARRESTING OFFICER AGENCY: Treasure Island PD
LAST AGENCY INSPECTION: 05/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 21:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: GRAY

DATE/TIME OF TEST: 06/17/2017 22:37:26

Diagnostics Check	OK	22:31
Air Blank	0.000	22:31
Control Test	0.081	22:31
Air Blank	0.000	22:32
Subject Sample # 1	0.000	22:32
Breath Volume	1.781	22:32
Air Blank	0.000	22:33
Air Blank	0.000	22:35
Subject Sample # 2	0.000	22:36
Breath Volume	1.718	22:36
Air Blank	0.000	22:36
Control Test	0.082	22:36
Air Blank	0.000	22:37
Diagnostics Check	OK	22:37

Data Download Date/Time: 06/28/2017 22:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ZARASVAND,J
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 22:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: MERRITT

DATE/TIME OF TEST: 06/17/2017 23:27:46

Diagnostics Check	OK	23:20
Air Blank	0.000	23:21
Control Test	0.081	23:21
Air Blank	0.000	23:22
Subject Sample # 1	0.264	23:22
Breath Volume	2.292	23:22
Air Blank	0.000	23:23
Air Blank	0.000	23:25
Subject Sample # 2	0.266	23:26
Breath Volume	2.394	23:26
Air Blank	0.000	23:26
Control Test	0.081	23:27
Air Blank	0.000	23:27
Diagnostics Check	OK	23:27

Data Download Date/Time: 06/28/2017 22:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TAVERNIER,M
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 00:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: OSHEA

DATE/TIME OF TEST: 06/18/2017 01:34:40

Diagnostics Check	OK	01:20	
Air Blank	0.000	01:21	
Control Test	0.082	01:21	
Air Blank	0.000	01:22	
Subject Sample	# 1 0.114	01:25	Volume Not Met
Breath Volume	0.671	01:25	
Air Blank	0.000	01:26	
Air Blank	0.000	01:27	
Subject Sample	# 2 0.132	01:29	
Breath Volume	2.468	01:29	
Air Blank	0.000	01:30	
Air Blank	0.000	01:32	
Subject Sample	# 3 0.136	01:33	
Breath Volume	1.871	01:33	
Air Blank	0.000	01:33	
Control Test	0.079	01:34	
Air Blank	0.000	01:34	
Diagnostics Check	OK	01:34	

Data Download Date/Time: 06/28/2017 22:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOOD,COREY,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HOOD,COREY
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 17:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:44

SUBJECT'S LAST NAME: BOYER

DATE/TIME OF TEST: 06/19/2017 19:13:23

Diagnostics Check	OK	19:07
Air Blank	0.000	19:07
Control Test	0.081	19:07
Air Blank	0.000	19:08
Subject Sample # 1	0.187	19:08
Breath Volume	1.730	19:08
Air Blank	0.000	19:09
Air Blank	0.000	19:11
Subject Sample # 2	0.192	19:11
Breath Volume	1.527	19:11
Air Blank	0.000	19:12
Control Test	0.081	19:12
Air Blank	0.000	19:13
Diagnostics Check	OK	19:13

Data Download Date/Time: 06/28/2017 22:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRIFFIN,KYLE,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TATE,GREGORY
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 04:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:55

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 06/20/2017 05:20:18

Diagnostics Check	OK	05:17	
Air Blank	0.000	05:17	
Control Test	0.082	05:17	
Air Blank	0.000	05:18	
Subject Sample	# 1	05:18	Subject Test Refused
Breath Volume	0.000	05:18	
Air Blank	0.000	05:19	
Control Test	0.082	05:19	
Air Blank	0.000	05:20	
Diagnostics Check	OK	05:20	

Data Download Date/Time: 06/28/2017 22:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRIFFIN,KYLE,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: PETERSON,WILLIAM
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 03:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: O'DONNELL

DATE/TIME OF TEST: 06/21/2017 04:30:31

Diagnostics Check	OK	04:26	
Air Blank	0.000	04:26	
Control Test	0.081	04:27	
Air Blank	0.000	04:27	
Subject Sample	# 1	04:29	Subject Test Refused
Breath Volume	0.000	04:29	
Air Blank	0.000	04:29	
Control Test	0.081	04:29	
Air Blank	0.000	04:30	
Diagnostics Check	OK	04:30	

Data Download Date/Time: 06/28/2017 22:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUARTE,CHRISTOPHER,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: DUARTE,CHRISTOPHER
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 19:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:11

SUBJECT'S LAST NAME: MEYERS

DATE/TIME OF TEST: 06/22/2017 20:42:12

Diagnostics Check	OK	20:36
Air Blank	0.000	20:36
Control Test	0.082	20:36
Air Blank	0.000	20:37
Subject Sample # 1	0.106	20:37
Breath Volume	1.984	20:37
Air Blank	0.000	20:38
Air Blank	0.000	20:40
Subject Sample # 2	0.105	20:40
Breath Volume	1.945	20:40
Air Blank	0.000	20:41
Control Test	0.082	20:41
Air Blank	0.000	20:42
Diagnostics Check	OK	20:42

Data Download Date/Time: 06/28/2017 22:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LANEY,DAMON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 00:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:34

SUBJECT'S LAST NAME: KROPINAK

DATE/TIME OF TEST: 06/23/2017 01:00:51

Diagnostics Check	OK	00:54
Air Blank	0.000	00:54
Control Test	0.081	00:55
Air Blank	0.000	00:55
Subject Sample # 1	0.107	00:56
Breath Volume	1.957	00:56
Air Blank	0.000	00:56
Air Blank	0.000	00:58
Subject Sample # 2	0.105	00:59
Breath Volume	1.425	00:59
Air Blank	0.000	00:59
Control Test	0.080	01:00
Air Blank	0.000	01:00
Diagnostics Check	OK	01:00

Data Download Date/Time: 06/28/2017 22:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,A,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ELLIS,A
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:27

SUBJECT'S LAST NAME: KORKAJ

DATE/TIME OF TEST: 06/23/2017 02:59:56

Diagnostics Check	OK	02:48	
Air Blank	0.000	02:48	
Control Test	0.082	02:49	
Air Blank	0.000	02:49	
Subject Sample	# 1	02:52	No Sample Provided
Breath Volume	0.000	02:52	
Air Blank	0.000	02:53	
Air Blank	0.000	02:55	
Subject Sample	# 2	0.000	02:55
Breath Volume	3.812	02:55	
Air Blank	0.000	02:56	
Air Blank	0.000	02:58	
Subject Sample	# 3	0.000	02:58
Breath Volume	2.941	02:58	
Air Blank	0.000	02:59	
Control Test	0.081	02:59	
Air Blank	0.000	02:59	
Diagnostics Check	OK	02:59	

Data Download Date/Time: 06/28/2017 22:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LANEY,DAMON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 04:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:49

SUBJECT'S LAST NAME: EVANS

DATE/TIME OF TEST: 06/23/2017 05:16:42

Diagnostics Check	OK	05:10
Air Blank	0.000	05:10
Control Test	0.082	05:11
Air Blank	0.000	05:11
Subject Sample # 1	0.000	05:12
Breath Volume	1.703	05:12
Air Blank	0.000	05:12
Air Blank	0.000	05:14
Subject Sample # 2	0.000	05:15
Breath Volume	1.636	05:15
Air Blank	0.000	05:15
Control Test	0.081	05:16
Air Blank	0.000	05:16
Diagnostics Check	OK	05:16

Data Download Date/Time: 06/28/2017 22:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: RICOTTILLI,J
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 23:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:22

SUBJECT'S LAST NAME: SCHAUER

DATE/TIME OF TEST: 06/24/2017 00:53:55

Diagnostics Check	OK	00:42	
Air Blank	0.000	00:42	
Control Test	0.081	00:42	
Air Blank	0.000	00:43	
Subject Sample	# 1 0.000	00:46	Volume Not Met
Breath Volume	0.015	00:46	
Air Blank	0.000	00:47	
Air Blank	0.000	00:48	
Subject Sample	# 2 0.197	00:52	Volume Not Met
Breath Volume	0.371	00:52	
Air Blank	0.000	00:52	
Control Test	0.079	00:53	
Air Blank	0.000	00:53	
Diagnostics Check	OK	00:53	

Data Download Date/Time: 06/28/2017 22:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: RICOTTILLI,J
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 23:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:22

SUBJECT'S LAST NAME: SCHAUER

DATE/TIME OF TEST: 06/24/2017 01:06:24

Diagnostics Check	OK	01:01	
Air Blank	0.000	01:01	
Control Test	0.078	01:02	
Air Blank	0.000	01:02	
Subject Sample	# 1	01:04	Subject Test Refused
Breath Volume	0.773	01:04	
Air Blank	0.000	01:05	
Control Test	0.080	01:05	
Air Blank	0.000	01:06	
Diagnostics Check	OK	01:06	

Data Download Date/Time: 06/28/2017 22:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: CAREY,K
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 23:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: VENTURA

DATE/TIME OF TEST: 06/24/2017 01:16:15

Diagnostics Check	OK	01:09
Air Blank	0.000	01:10
Control Test	0.078	01:10
Air Blank	0.000	01:10
Subject Sample # 1	0.072	01:11
Breath Volume	1.230	01:11
Air Blank	0.000	01:11
Air Blank	0.000	01:13
Subject Sample # 2	0.080	01:14
Breath Volume	1.816	01:14
Air Blank	0.000	01:15
Control Test	0.079	01:15
Air Blank	0.000	01:16
Diagnostics Check	OK	01:16

Data Download Date/Time: 06/28/2017 22:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,A,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ELLIS,A
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 03:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: SALAZAR

DATE/TIME OF TEST: 06/24/2017 04:14:14

Diagnostics Check	OK	04:11	
Air Blank	0.000	04:11	
Control Test	0.081	04:11	
Air Blank	0.000	04:12	
Subject Sample	# 1	04:12	Subject Test Refused
Breath Volume	0.000	04:12	
Air Blank	0.000	04:13	
Control Test	0.081	04:13	
Air Blank	0.000	04:14	
Diagnostics Check	OK	04:14	

Data Download Date/Time: 06/28/2017 22:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LANEY,DAMON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 05:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:08

SUBJECT'S LAST NAME: TUSHAUS

DATE/TIME OF TEST: 06/24/2017 05:32:37

Diagnostics Check	OK	05:29	
Air Blank	0.000	05:30	
Control Test	0.081	05:30	
Air Blank	0.000	05:30	
Subject Sample	# 1	05:31	Subject Test Refused
Breath Volume	0.000	05:31	
Air Blank	0.000	05:31	
Control Test	0.081	05:32	
Air Blank	0.000	05:32	
Diagnostics Check	OK	05:32	

Data Download Date/Time: 06/28/2017 22:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,A,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ELLIS,A
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 23:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: NEAL

DATE/TIME OF TEST: 06/25/2017 00:05:38

Diagnostics Check	OK	00:02	
Air Blank	0.000	00:02	
Control Test	0.081	00:03	
Air Blank	0.000	00:03	
Subject Sample	# 1	00:04	Subject Test Refused
Breath Volume	0.000	00:04	
Air Blank	0.000	00:04	
Control Test	0.080	00:05	
Air Blank	0.000	00:05	
Diagnostics Check	OK	00:05	

Data Download Date/Time: 06/28/2017 22:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LANEY,DAMON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 01:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:14

SUBJECT'S LAST NAME: RUMSEY

DATE/TIME OF TEST: 06/25/2017 01:40:24

Diagnostics Check	OK	01:34
Air Blank	0.000	01:34
Control Test	0.082	01:34
Air Blank	0.000	01:35
Subject Sample # 1	0.041	01:36
Breath Volume	1.441	01:36
Air Blank	0.000	01:36
Air Blank	0.000	01:38
Subject Sample # 2	0.043	01:38
Breath Volume	1.824	01:38
Air Blank	0.000	01:39
Control Test	0.081	01:39
Air Blank	0.000	01:40
Diagnostics Check	OK	01:40

Data Download Date/Time: 06/28/2017 22:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH, WILLIAM, T
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SMITH, WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 03:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:08

SUBJECT'S LAST NAME: GERMOSEN

DATE/TIME OF TEST: 06/29/2017 04:36:11

Diagnostics Check	OK	04:28
Air Blank	0.000	04:28
Control Test	0.082	04:29
Air Blank	0.000	04:29
Subject Sample # 1	0.110	04:30
Breath Volume	1.312	04:30
Air Blank	0.000	04:31
Air Blank	0.000	04:33
Subject Sample # 2	0.108	04:34
Breath Volume	1.617	04:34
Air Blank	0.000	04:35
Control Test	0.081	04:35
Air Blank	0.000	04:36
Diagnostics Check	OK	04:36

Data Download Date/Time: 07/28/2017 22:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LANGE,K
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: SAGNELLA

DATE/TIME OF TEST: 06/30/2017 03:53:36

Diagnostics Check	OK	03:50	
Air Blank	0.000	03:51	
Control Test	0.081	03:51	
Air Blank	0.000	03:52	
Subject Sample	# 1	03:52	Subject Test Refused
Breath Volume	0.000	03:52	
Air Blank	0.000	03:52	
Control Test	0.081	03:53	
Air Blank	0.000	03:53	
Diagnostics Check	OK	03:53	

Data Download Date/Time: 07/28/2017 22:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005337

INSTRUMENT REGISTERED TO: PERRY POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SWINDLE, CHARLES, M
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: SWINDLE, CHARLES
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 05/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG619605
DRY GAS EXPIRY DATE: 07/14/2018
VIOLATION CODE: DUI
ARREST TIME: 16:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:50

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 06/04/2017 17:32:42

Diagnostics Check	OK	17:26
Air Blank	0.000	17:27
Control Test	0.080	17:27
Air Blank	0.000	17:27
Subject Sample # 1	0.044	17:28
Breath Volume	1.605	17:28
Air Blank	0.000	17:28
Air Blank	0.000	17:30
Subject Sample # 2	0.044	17:31
Breath Volume	1.191	17:31
Air Blank	0.000	17:31
Control Test	0.080	17:32
Air Blank	0.000	17:32
Diagnostics Check	OK	17:32

Data Download Date/Time: 06/29/2017 15:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COX,LENARD,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ALLI,ANTHONY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A2
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:39

SUBJECT'S LAST NAME: FERRER

DATE/TIME OF TEST: 06/15/2017 02:26:44

Diagnostics Check	OK	02:20
Air Blank	0.000	02:20
Control Test	0.078	02:21
Air Blank	0.000	02:21
Subject Sample # 1	0.242	02:22
Breath Volume	3.242	02:22
Air Blank	0.000	02:22
Air Blank	0.000	02:24
Subject Sample # 2	0.230	02:25
Breath Volume	3.078	02:25
Air Blank	0.000	02:25
Control Test	0.078	02:26
Air Blank	0.000	02:26
Diagnostics Check	OK	02:26

Data Download Date/Time: 07/04/2017 16:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COX,LENARD,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: KARAYIANES,MICHAEL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A2
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:21

SUBJECT'S LAST NAME: MATHUS

DATE/TIME OF TEST: 06/15/2017 03:52:36

Diagnostics Check	OK	03:46
Air Blank	0.000	03:46
Control Test	0.078	03:47
Air Blank	0.000	03:47
Subject Sample # 1	0.000	03:48
Breath Volume	3.082	03:48
Air Blank	0.000	03:48
Air Blank	0.000	03:50
Subject Sample # 2	0.000	03:51
Breath Volume	3.156	03:51
Air Blank	0.000	03:51
Control Test	0.078	03:52
Air Blank	0.000	03:52
Diagnostics Check	OK	03:52

Data Download Date/Time: 07/04/2017 16:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KARAYIANES,MICHAEL,F
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: COX,LENARD
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A2
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 06/16/2017 00:22:45

Diagnostics Check	OK	00:16
Air Blank	0.000	00:17
Control Test	0.077	00:17
Air Blank	0.000	00:17
Subject Sample # 1	0.000	00:18
Breath Volume	3.511	00:18
Air Blank	0.000	00:18
Air Blank	0.000	00:20
Subject Sample # 2	0.000	00:21
Breath Volume	3.453	00:21
Air Blank	0.000	00:21
Control Test	0.075	00:22
Air Blank	0.000	00:22
Diagnostics Check	OK	00:22

Data Download Date/Time: 07/04/2017 16:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLI, ANTHONY,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: COX, LENARD
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A2
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 19:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:45

SUBJECT'S LAST NAME: SALAMONIK

DATE/TIME OF TEST: 06/21/2017 19:15:10

Diagnostics Check	OK	19:06	
Air Blank	0.000	19:06	
Control Test	0.077	19:06	
Air Blank	0.000	19:07	
Subject Sample	# 1 0.188	19:10	Volume Not Met
Breath Volume	1.031	19:10	
Air Blank	0.000	19:11	
Air Blank	0.000	19:12	
Subject Sample	# 2	19:13	Subject Test Refused
Breath Volume	0.000	19:13	
Air Blank	0.000	19:14	
Control Test	0.077	19:14	
Air Blank	0.000	19:15	
Diagnostics Check	OK	19:15	

Data Download Date/Time: 07/04/2017 16:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COX,LENARD,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ALLI,ANTHONY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A2
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:16

SUBJECT'S LAST NAME: DEES

DATE/TIME OF TEST: 06/21/2017 21:44:32

Diagnostics Check	OK	21:41	
Air Blank	0.000	21:41	
Control Test	0.077	21:42	
Air Blank	0.000	21:42	
Subject Sample	# 1	21:43	RFI Detect
Breath Volume	0.000	21:43	
Air Blank		21:44	RFI Detect
Air Blank		21:44	RFI Detect

Data Download Date/Time: 07/04/2017 16:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COX,LENARD,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ALLI,ANTHONY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A2
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:16

SUBJECT'S LAST NAME: DEES

DATE/TIME OF TEST: 06/21/2017 21:54:45

Diagnostics Check	OK	21:46	
Air Blank	0.000	21:46	
Control Test	0.078	21:46	
Air Blank	0.000	21:47	
Subject Sample	# 1 0.000	21:50	Volume Not Met
Breath Volume	0.054	21:50	
Air Blank	0.000	21:51	
Air Blank	0.000	21:53	
Subject Sample	# 2	21:53	Subject Test Refused
Breath Volume	0.000	21:53	
Air Blank	0.000	21:53	
Control Test	0.077	21:54	
Air Blank	0.000	21:54	
Diagnostics Check	OK	21:54	

Data Download Date/Time: 07/04/2017 16:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COX,LENARD,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: WEISKOPF,MICHAEL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A2
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:08

SUBJECT'S LAST NAME: DAYTON

DATE/TIME OF TEST: 06/22/2017 02:44:11

Diagnostics Check	OK	02:38	
Air Blank	0.000	02:38	
Control Test	0.078	02:38	
Air Blank	0.000	02:39	
Subject Sample	# 1 0.207	02:40	
Breath Volume	2.687	02:40	
Air Blank	0.000	02:40	
Air Blank	0.000	02:42	
Subject Sample	# 2	02:42	RFI Detect
Breath Volume	0.000	02:42	
Air Blank	0.000	02:43	
Control Test	0.077	02:43	
Air Blank	0.000	02:44	
Diagnostics Check	OK	02:44	

Data Download Date/Time: 07/04/2017 16:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COX,LENARD,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: WEISKOPF,MICHAEL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A2
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:08

SUBJECT'S LAST NAME: DAYTON

DATE/TIME OF TEST: 06/22/2017 02:52:19

Diagnostics Check	OK	02:45
Air Blank	0.000	02:46
Control Test	0.078	02:46
Air Blank	0.000	02:47
Subject Sample # 1	0.193	02:47
Breath Volume	2.011	02:47
Air Blank	0.000	02:48
Air Blank	0.000	02:50
Subject Sample # 2	0.186	02:50
Breath Volume	1.328	02:50
Air Blank	0.000	02:51
Control Test	0.076	02:51
Air Blank	0.000	02:52
Diagnostics Check	OK	02:52

Data Download Date/Time: 07/04/2017 16:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KARAYIANES,MICHAEL,F
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: COX,LENARD
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A2
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 06/25/2017 00:12:33

Diagnostics Check	OK	00:06
Air Blank	0.000	00:06
Control Test	0.077	00:06
Air Blank	0.000	00:07
Subject Sample # 1	0.257	00:07
Breath Volume	2.523	00:07
Air Blank	0.000	00:08
Air Blank	0.000	00:10
Subject Sample # 2	0.247	00:11
Breath Volume	2.664	00:11
Air Blank	0.000	00:11
Control Test	0.076	00:11
Air Blank	0.000	00:12
Diagnostics Check	OK	00:12

Data Download Date/Time: 07/04/2017 16:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COX,LENARD,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ZOLLOTA,VLADIMIR
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A2
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: SHUSHPANOVA

DATE/TIME OF TEST: 06/29/2017 00:13:30

Diagnostics Check	OK	00:10	
Air Blank	0.000	00:10	
Control Test	0.076	00:11	
Air Blank	0.000	00:11	
Subject Sample	# 1	00:12	Subject Test Refused
Breath Volume	0.000	00:12	
Air Blank	0.000	00:12	
Control Test	0.077	00:12	
Air Blank	0.000	00:13	
Diagnostics Check	OK	00:13	

Data Download Date/Time: 07/04/2017 16:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005402

INSTRUMENT REGISTERED TO: PORT ORANGE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SYDESKI, BLAKE, S
BREATH TEST OPERATOR AGENCY: Port Orange PD
ARRESTING OFFICER: SYDESKI, BLAKE
ARRESTING OFFICER AGENCY: Port Orange PD
LAST AGENCY INSPECTION: 05/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1836890
DRY GAS EXPIRY DATE: 03/26/2018
VIOLATION CODE: DUI
ARREST TIME: 01:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:56

SUBJECT'S LAST NAME: EPPS

DATE/TIME OF TEST: 06/03/2017 02:23:21

Diagnostics Check	OK	02:16
Air Blank	0.000	02:17
Control Test	0.081	02:17
Air Blank	0.000	02:18
Subject Sample # 1	0.179	02:18
Breath Volume	2.496	02:18
Air Blank	0.000	02:19
Air Blank	0.000	02:21
Subject Sample # 2	0.183	02:21
Breath Volume	2.355	02:21
Air Blank	0.000	02:22
Control Test	0.081	02:22
Air Blank	0.000	02:23
Diagnostics Check	OK	02:23

Data Download Date/Time: 06/30/2017 14:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005402

INSTRUMENT REGISTERED TO: PORT ORANGE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRISON, WILLIAM, C
BREATH TEST OPERATOR AGENCY: Port Orange PD
ARRESTING OFFICER: N/A, N/A
ARRESTING OFFICER AGENCY: Port Orange PD
LAST AGENCY INSPECTION: 05/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1836890
DRY GAS EXPIRY DATE: 03/26/2018
VIOLATION CODE: ADMIN
ARREST TIME: 12:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:45

SUBJECT'S LAST NAME: AQUINO

DATE/TIME OF TEST: 06/08/2017 14:20:45

Diagnostics Check	OK	14:14
Air Blank	0.000	14:15
Control Test	0.082	14:15
Air Blank	0.000	14:15
Subject Sample # 1	0.000	14:16
Breath Volume	3.050	14:16
Air Blank	0.000	14:16
Air Blank	0.000	14:18
Subject Sample # 2	0.000	14:19
Breath Volume	3.117	14:19
Air Blank	0.000	14:19
Control Test	0.082	14:20
Air Blank	0.000	14:20
Diagnostics Check	OK	14:20

Data Download Date/Time: 06/30/2017 14:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005402

INSTRUMENT REGISTERED TO: PORT ORANGE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SYDESKI, BLAKE, S
BREATH TEST OPERATOR AGENCY: Port Orange PD
ARRESTING OFFICER: FISCHETTI, JAMES
ARRESTING OFFICER AGENCY: Port Orange PD
LAST AGENCY INSPECTION: 05/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1836890
DRY GAS EXPIRY DATE: 03/26/2018
VIOLATION CODE: DUI
ARREST TIME: 05:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:45

SUBJECT'S LAST NAME: HOLT

DATE/TIME OF TEST: 06/11/2017 06:15:53

Diagnostics Check	OK	06:08
Air Blank	0.000	06:08
Control Test	0.081	06:08
Air Blank	0.000	06:09
Subject Sample # 1	0.156	06:11
Breath Volume	1.242	06:11
Air Blank	0.000	06:11
Air Blank	0.000	06:13
Subject Sample # 2	0.158	06:14
Breath Volume	1.414	06:14
Air Blank	0.000	06:14
Control Test	0.081	06:15
Air Blank	0.000	06:15
Diagnostics Check	OK	06:15

Data Download Date/Time: 06/30/2017 14:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005402

INSTRUMENT REGISTERED TO: PORT ORANGE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIDONATO,MICHAEL,J
BREATH TEST OPERATOR AGENCY: South Daytona PD
ARRESTING OFFICER: DIDONATO,MICHAEL
ARRESTING OFFICER AGENCY: South Daytona PD
LAST AGENCY INSPECTION: 05/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1836890
DRY GAS EXPIRY DATE: 03/26/2018
VIOLATION CODE: DUI
ARREST TIME: 22:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: RAYMOND

DATE/TIME OF TEST: 06/13/2017 23:48:30

Diagnostics Check	OK	23:42
Air Blank	0.000	23:42
Control Test	0.082	23:42
Air Blank	0.000	23:43
Subject Sample # 1	0.118	23:44
Breath Volume	1.281	23:44
Air Blank	0.000	23:44
Air Blank	0.000	23:46
Subject Sample # 2	0.122	23:46
Breath Volume	1.601	23:46
Air Blank	0.000	23:47
Control Test	0.083	23:47
Air Blank	0.000	23:48
Diagnostics Check	OK	23:48

Data Download Date/Time: 06/30/2017 14:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005402

INSTRUMENT REGISTERED TO: PORT ORANGE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRISON, WILLIAM, C
BREATH TEST OPERATOR AGENCY: Port Orange PD
ARRESTING OFFICER: N/A, N/A
ARRESTING OFFICER AGENCY: Port Orange PD
LAST AGENCY INSPECTION: 05/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1836890
DRY GAS EXPIRY DATE: 03/26/2018
VIOLATION CODE: ADMIN
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:30

SUBJECT'S LAST NAME: COWAN

DATE/TIME OF TEST: 06/20/2017 09:56:57

Diagnostics Check	OK	09:50
Air Blank	0.000	09:51
Control Test	0.082	09:51
Air Blank	0.000	09:52
Subject Sample # 1	0.000	09:52
Breath Volume	3.179	09:52
Air Blank	0.000	09:53
Air Blank	0.000	09:55
Subject Sample # 2	0.000	09:55
Breath Volume	1.972	09:55
Air Blank	0.000	09:56
Control Test	0.082	09:56
Air Blank	0.000	09:56
Diagnostics Check	OK	09:56

Data Download Date/Time: 06/30/2017 14:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005402

INSTRUMENT REGISTERED TO: PORT ORANGE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MANDESE,ANGELO,
BREATH TEST OPERATOR AGENCY: Port Orange PD
ARRESTING OFFICER: DAVIS,JASON
ARRESTING OFFICER AGENCY: Port Orange PD
LAST AGENCY INSPECTION: 05/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1836890
DRY GAS EXPIRY DATE: 03/26/2018
VIOLATION CODE: DUI
ARREST TIME: 22:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: INGRAM

DATE/TIME OF TEST: 06/22/2017 23:09:38

Diagnostics Check	OK	23:03
Air Blank	0.000	23:03
Control Test	0.081	23:04
Air Blank	0.000	23:04
Subject Sample # 1	0.130	23:05
Breath Volume	2.105	23:05
Air Blank	0.000	23:05
Air Blank	0.000	23:07
Subject Sample # 2	0.131	23:08
Breath Volume	1.398	23:08
Air Blank	0.000	23:08
Control Test	0.083	23:09
Air Blank	0.000	23:09
Diagnostics Check	OK	23:09

Data Download Date/Time: 06/30/2017 14:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005402

INSTRUMENT REGISTERED TO: PORT ORANGE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BINES,BRANDON,W
BREATH TEST OPERATOR AGENCY: Port Orange PD
ARRESTING OFFICER: BINES,BRANDON
ARRESTING OFFICER AGENCY: Ponce Inlet PD
LAST AGENCY INSPECTION: 05/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1836890
DRY GAS EXPIRY DATE: 03/26/2018
VIOLATION CODE: DUI
ARREST TIME: 14:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:15

SUBJECT'S LAST NAME: SCHAIER

DATE/TIME OF TEST: 06/25/2017 15:51:43

Diagnostics Check	OK	15:36	
Air Blank	0.000	15:36	
Control Test	0.082	15:36	
Air Blank	0.000	15:37	
Subject Sample	# 1 0.072	15:40	Volume Not Met
Breath Volume	0.777	15:40	
Air Blank	0.000	15:41	
Air Blank	0.000	15:42	
Subject Sample	# 2 0.070	15:44	
Breath Volume	1.738	15:44	
Air Blank	0.000	15:45	
Air Blank	0.000	15:47	
Subject Sample	# 3 0.069	15:50	Volume Not Met
Breath Volume	0.207	15:50	
Air Blank	0.000	15:50	
Control Test	0.079	15:51	
Air Blank	0.000	15:51	
Diagnostics Check	OK	15:51	

Data Download Date/Time: 06/30/2017 14:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005402

INSTRUMENT REGISTERED TO: PORT ORANGE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BENEZETTE,BENJAMIN,J
BREATH TEST OPERATOR AGENCY: Port Orange PD
ARRESTING OFFICER: ZEPHIR,JEFF
ARRESTING OFFICER AGENCY: Daytona Beach Shores
LAST AGENCY INSPECTION: 05/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1836890
DRY GAS EXPIRY DATE: 03/26/2018
VIOLATION CODE: DUI
ARREST TIME: 00:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:44

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 06/28/2017 02:16:25

Diagnostics Check	OK	02:10
Air Blank	0.000	02:10
Control Test	0.083	02:10
Air Blank	0.000	02:11
Subject Sample # 1	0.133	02:11
Breath Volume	2.386	02:11
Air Blank	0.000	02:12
Air Blank	0.000	02:14
Subject Sample # 2	0.129	02:14
Breath Volume	2.031	02:14
Air Blank	0.000	02:15
Control Test	0.083	02:15
Air Blank	0.000	02:16
Diagnostics Check	OK	02:16

Data Download Date/Time: 06/30/2017 14:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005454

INSTRUMENT REGISTERED TO: OCALA POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOEDICKE, WILLIAM, R
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: JOEDICKE, WILLIAM
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 05/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: BAXTER

DATE/TIME OF TEST: 06/02/2017 00:26:31

Diagnostics Check	OK	00:19
Air Blank	0.000	00:20
Control Test	0.077	00:20
Air Blank	0.000	00:21
Subject Sample # 1	0.099	00:21
Breath Volume	2.937	00:21
Air Blank	0.000	00:22
Air Blank	0.000	00:24
Subject Sample # 2	0.094	00:24
Breath Volume	3.132	00:24
Air Blank	0.000	00:25
Control Test	0.078	00:25
Air Blank	0.000	00:26
Diagnostics Check	OK	00:26

Data Download Date/Time: 06/29/2017 23:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005454

INSTRUMENT REGISTERED TO: OCALA POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOEDICKE, WILLIAM, R
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: JOEDICKE, WILLIAM
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 05/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: NETTLES

DATE/TIME OF TEST: 06/02/2017 22:14:06

Diagnostics Check	OK	22:06
Air Blank	0.000	22:06
Control Test	0.077	22:07
Air Blank	0.000	22:07
Subject Sample # 1	0.146	22:09
Breath Volume	1.109	22:09
Air Blank	0.000	22:10
Air Blank	0.000	22:11
Subject Sample # 2	0.160	22:12
Breath Volume	1.183	22:12
Air Blank	0.000	22:13
Control Test	0.077	22:13
Air Blank	0.000	22:14
Diagnostics Check	OK	22:14

Data Download Date/Time: 06/29/2017 23:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005454

INSTRUMENT REGISTERED TO: OCALA POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOEDICKE, WILLIAM, R
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: JOEDICKE, WILLIAM
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 05/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: ELIZONDO

DATE/TIME OF TEST: 06/07/2017 02:42:27

Diagnostics Check	OK	02:36
Air Blank	0.000	02:36
Control Test	0.077	02:36
Air Blank	0.000	02:37
Subject Sample # 1	0.060	02:37
Breath Volume	3.007	02:37
Air Blank	0.000	02:38
Air Blank	0.000	02:40
Subject Sample # 2	0.060	02:40
Breath Volume	3.554	02:40
Air Blank	0.000	02:41
Control Test	0.077	02:41
Air Blank	0.000	02:42
Diagnostics Check	OK	02:42

Data Download Date/Time: 06/29/2017 23:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005454

INSTRUMENT REGISTERED TO: OCALA POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOEDICKE, WILLIAM, R
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: JOEDICKE, WILLIAM
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 05/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: COLLYER

DATE/TIME OF TEST: 06/08/2017 01:51:10

Diagnostics Check	OK	01:44
Air Blank	0.000	01:45
Control Test	0.079	01:45
Air Blank	0.000	01:46
Subject Sample # 1	0.116	01:46
Breath Volume	1.238	01:46
Air Blank	0.000	01:47
Air Blank	0.000	01:48
Subject Sample # 2	0.113	01:49
Breath Volume	1.886	01:49
Air Blank	0.000	01:50
Control Test	0.078	01:50
Air Blank	0.000	01:51
Diagnostics Check	OK	01:51

Data Download Date/Time: 06/29/2017 23:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005454

INSTRUMENT REGISTERED TO: OCALA POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOEDICKE, WILLIAM, R
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: JOEDICKE, WILLIAM
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 05/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:22

SUBJECT'S LAST NAME: BRECHT

DATE/TIME OF TEST: 06/08/2017 02:01:34

Diagnostics Check	OK	01:55
Air Blank	0.000	01:55
Control Test	0.077	01:56
Air Blank	0.000	01:56
Subject Sample # 1	0.198	01:57
Breath Volume	3.828	01:57
Air Blank	0.000	01:57
Air Blank	0.000	01:59
Subject Sample # 2	0.193	01:59
Breath Volume	2.968	01:59
Air Blank	0.000	02:00
Control Test	0.077	02:00
Air Blank	0.000	02:01
Diagnostics Check	OK	02:01

Data Download Date/Time: 06/29/2017 23:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005454

INSTRUMENT REGISTERED TO: OCALA POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,GARY,
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: MILLER,GARY
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 05/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: REYES

DATE/TIME OF TEST: 06/09/2017 04:06:00

Diagnostics Check	OK	03:59
Air Blank	0.000	04:00
Control Test	0.079	04:00
Air Blank	0.000	04:00
Subject Sample # 1	0.156	04:01
Breath Volume	2.339	04:01
Air Blank	0.000	04:02
Air Blank	0.000	04:04
Subject Sample # 2	0.161	04:04
Breath Volume	2.230	04:04
Air Blank	0.000	04:05
Control Test	0.078	04:05
Air Blank	0.000	04:05
Diagnostics Check	OK	04:06

Data Download Date/Time: 06/29/2017 23:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005454

INSTRUMENT REGISTERED TO: OCALA POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOEDICKE, WILLIAM, R
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: JOEDICKE, WILLIAM
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 05/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: BISHOP

DATE/TIME OF TEST: 06/10/2017 03:11:41

Diagnostics Check	OK	03:04
Air Blank	0.000	03:05
Control Test	0.079	03:05
Air Blank	0.000	03:06
Subject Sample # 1	0.219	03:06
Breath Volume	1.488	03:06
Air Blank	0.000	03:07
Air Blank	0.000	03:09
Subject Sample # 2	0.220	03:10
Breath Volume	1.910	03:10
Air Blank	0.000	03:10
Control Test	0.078	03:11
Air Blank	0.000	03:11
Diagnostics Check	OK	03:11

Data Download Date/Time: 06/29/2017 23:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005454

INSTRUMENT REGISTERED TO: OCALA POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOEDICKE, WILLIAM, R
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: JOEDICKE, WILLIAM
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 05/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:58

SUBJECT'S LAST NAME: LUNDY

DATE/TIME OF TEST: 06/21/2017 19:24:48

Diagnostics Check	OK	19:18
Air Blank	0.000	19:18
Control Test	0.076	19:19
Air Blank	0.000	19:19
Subject Sample # 1	0.118	19:20
Breath Volume	1.457	19:20
Air Blank	0.000	19:20
Air Blank	0.000	19:22
Subject Sample # 2	0.120	19:23
Breath Volume	1.132	19:23
Air Blank	0.000	19:23
Control Test	0.077	19:24
Air Blank	0.000	19:24
Diagnostics Check	OK	19:24

Data Download Date/Time: 06/29/2017 23:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005454

INSTRUMENT REGISTERED TO: OCALA POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,GARY,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: MILLER,GARY
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 05/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:32

SUBJECT'S LAST NAME: VALLEJO CORTEZ

DATE/TIME OF TEST: 06/22/2017 00:16:24

Diagnostics Check	OK	00:10
Air Blank	0.000	00:10
Control Test	0.077	00:10
Air Blank	0.000	00:11
Subject Sample # 1	0.096	00:11
Breath Volume	3.593	00:11
Air Blank	0.000	00:12
Air Blank	0.000	00:14
Subject Sample # 2	0.092	00:14
Breath Volume	2.542	00:14
Air Blank	0.000	00:15
Control Test	0.077	00:15
Air Blank	0.000	00:16
Diagnostics Check	OK	00:16

Data Download Date/Time: 06/29/2017 23:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005454

INSTRUMENT REGISTERED TO: OCALA POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOEDICKE, WILLIAM, R
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: JOEDICKE, WILLIAM
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 05/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:23

SUBJECT'S LAST NAME: BERMUDEZ

DATE/TIME OF TEST: 06/23/2017 01:50:00

Diagnostics Check	OK	01:43
Air Blank	0.000	01:44
Control Test	0.077	01:44
Air Blank	0.000	01:44
Subject Sample # 1	0.065	01:45
Breath Volume	2.468	01:45
Air Blank	0.000	01:46
Air Blank	0.000	01:47
Subject Sample # 2	0.066	01:48
Breath Volume	2.890	01:48
Air Blank	0.000	01:49
Control Test	0.077	01:49
Air Blank	0.000	01:49
Diagnostics Check	OK	01:50

Data Download Date/Time: 06/29/2017 23:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005454

INSTRUMENT REGISTERED TO: OCALA POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOEDICKE, WILLIAM, R
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: JOEDICKE, WILLIAM
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 05/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:43

SUBJECT'S LAST NAME: MARTELLI

DATE/TIME OF TEST: 06/23/2017 22:13:07

Diagnostics Check	OK	22:06
Air Blank	0.000	22:07
Control Test	0.077	22:07
Air Blank	0.000	22:08
Subject Sample # 1	0.111	22:08
Breath Volume	4.101	22:08
Air Blank	0.000	22:09
Air Blank	0.000	22:10
Subject Sample # 2	0.115	22:11
Breath Volume	4.441	22:11
Air Blank	0.000	22:12
Control Test	0.077	22:12
Air Blank	0.000	22:13
Diagnostics Check	OK	22:13

Data Download Date/Time: 06/29/2017 23:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLAFLIN, DAVID, S
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: GOULD, LEONARD
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29614080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 23:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:04

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 06/03/2017 00:32:25

Diagnostics Check	OK	00:26	
Air Blank	0.000	00:26	
Control Test	0.081	00:26	
Air Blank	0.000	00:27	
Subject Sample # 1	0.278	00:28	
Breath Volume	2.953	00:28	
Air Blank	0.000	00:29	
Air Blank	0.000	00:30	
Subject Sample # 2	0.285	00:31	
Breath Volume	2.632	00:31	
Air Blank		00:31	Purge Fail
Air Blank	0.000	00:32	

Data Download Date/Time: 07/26/2017 10:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLAFLIN, DAVID, S
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: GOULD, LEONARD
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29614080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 23:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:04

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 06/03/2017 00:42:12

Diagnostics Check	OK	00:35
Air Blank	0.000	00:35
Control Test	0.081	00:35
Air Blank	0.000	00:36
Subject Sample # 1	0.265	00:37
Breath Volume	2.773	00:37
Air Blank	0.000	00:38
Air Blank	0.000	00:39
Subject Sample # 2	0.256	00:40
Breath Volume	2.722	00:40
Air Blank	0.000	00:41
Control Test	0.082	00:41
Air Blank	0.000	00:42
Diagnostics Check	OK	00:42

Data Download Date/Time: 07/26/2017 10:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLAFLIN,DAVID,S
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: ELLIS,BILLY
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29614080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 00:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 06/03/2017 01:28:44

Diagnostics Check	OK	01:27	
Air Blank		01:28	Ambient Fail
Air Blank	0.000	01:28	

Data Download Date/Time: 07/26/2017 10:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLAFLIN,DAVID,S
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: ELLIS,BILLY
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29614080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 00:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 06/03/2017 01:37:58

Diagnostics Check	OK	01:30	
Air Blank	0.000	01:31	
Control Test	0.081	01:31	
Air Blank	0.000	01:31	
Subject Sample	# 1 0.218	01:35	Volume Not Met
Breath Volume	0.980	01:35	
Air Blank	0.000	01:36	
Air Blank		01:37	Ambient Fail
Air Blank	0.000	01:37	

Data Download Date/Time: 07/26/2017 10:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLAFLIN, DAVID, S
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: ELLIS, BILLY
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29614080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 00:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 06/03/2017 01:46:58

Diagnostics Check	OK	01:39	
Air Blank	0.000	01:39	
Control Test	0.080	01:40	
Air Blank	0.000	01:40	
Subject Sample # 1	0.247	01:41	
Breath Volume	1.632	01:41	
Air Blank	0.000	01:42	
Air Blank	0.000	01:43	
Subject Sample # 2	0.236	01:45	
Breath Volume	1.152	01:45	
Air Blank		01:46	Purge Fail
Air Blank	0.000	01:46	

Data Download Date/Time: 07/26/2017 10:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005477

INSTRUMENT REGISTERED TO: DADE CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BALDWIN,DEAN,
BREATH TEST OPERATOR AGENCY: Dade City PD
ARRESTING OFFICER: BELL,C
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 05/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A6
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 09:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:50

SUBJECT'S LAST NAME: HOFFMAN

DATE/TIME OF TEST: 06/13/2017 11:50:27

Diagnostics Check	OK	11:37
Air Blank	0.000	11:38
Control Test	0.079	11:38
Air Blank	0.000	11:39
Subject Sample # 1	0.255	11:39
Breath Volume	1.425	11:39
Air Blank	0.000	11:40
Air Blank	0.000	11:42
Subject Sample # 2	0.234	11:45
Breath Volume	2.328	11:45
Air Blank	0.000	11:46
Air Blank	0.000	11:47
Subject Sample # 3	0.251	11:48
Breath Volume	2.386	11:48
Air Blank	0.000	11:49
Control Test	0.075	11:49
Air Blank	0.000	11:50
Diagnostics Check	OK	11:50

Data Download Date/Time: 06/13/2017 15:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005537

INSTRUMENT REGISTERED TO: DAYTONA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KILKER,JASON,A
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: WALLIS,R
ARRESTING OFFICER AGENCY: Holly Hills PD
LAST AGENCY INSPECTION: 05/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14968
DRY GAS EXPIRY DATE: 08/06/2018
VIOLATION CODE: DUI
ARREST TIME: 11:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:20

SUBJECT'S LAST NAME: MELLEN

DATE/TIME OF TEST: 06/11/2017 12:49:09

Diagnostics Check	OK	12:41
Air Blank	0.000	12:42
Control Test	0.080	12:42
Air Blank	0.000	12:43
Subject Sample # 1	0.133	12:44
Breath Volume	1.546	12:44
Air Blank	0.000	12:45
Air Blank	0.000	12:46
Subject Sample # 2	0.143	12:47
Breath Volume	2.179	12:47
Air Blank	0.000	12:48
Control Test	0.078	12:48
Air Blank	0.000	12:49
Diagnostics Check	OK	12:49

Data Download Date/Time: 06/26/2017 15:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005537

INSTRUMENT REGISTERED TO: DAYTONA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KILKER,J,A
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: CYRUS,J
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 05/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14968
DRY GAS EXPIRY DATE: 08/06/2018
VIOLATION CODE: DUI
ARREST TIME: 06:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:42

SUBJECT'S LAST NAME: GILBERT

DATE/TIME OF TEST: 06/14/2017 07:10:14

Diagnostics Check	OK	07:03
Air Blank	0.000	07:04
Control Test	0.080	07:04
Air Blank	0.000	07:05
Subject Sample # 1	0.272	07:05
Breath Volume	2.375	07:05
Air Blank	0.000	07:06
Air Blank	0.000	07:07
Subject Sample # 2	0.275	07:08
Breath Volume	2.558	07:08
Air Blank	0.000	07:09
Control Test	0.079	07:09
Air Blank	0.000	07:10
Diagnostics Check	OK	07:10

Data Download Date/Time: 06/26/2017 15:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005537

INSTRUMENT REGISTERED TO: DAYTONA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAMPE,CORTLAND,B
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: EDWARDS,L
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 05/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14968
DRY GAS EXPIRY DATE: 08/06/2018
VIOLATION CODE: DUI
ARREST TIME: 03:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: DOCHTERMAN

DATE/TIME OF TEST: 06/17/2017 04:58:21

Diagnostics Check	OK	04:51
Air Blank	0.000	04:51
Control Test	0.080	04:51
Air Blank	0.000	04:52
Subject Sample # 1	0.176	04:53
Breath Volume	2.792	04:53
Air Blank	0.000	04:54
Air Blank	0.000	04:55
Subject Sample # 2	0.170	04:56
Breath Volume	2.382	04:56
Air Blank	0.000	04:57
Control Test	0.080	04:57
Air Blank	0.000	04:58
Diagnostics Check	OK	04:58

Data Download Date/Time: 06/26/2017 15:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005537

INSTRUMENT REGISTERED TO: DAYTONA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BISSA,JAMES,D
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: BISSA,J
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 05/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14968
DRY GAS EXPIRY DATE: 08/06/2018
VIOLATION CODE: ADMIN
ARREST TIME: 14:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:50

SUBJECT'S LAST NAME: TEST

DATE/TIME OF TEST: 06/23/2017 15:19:35

Diagnostics Check	OK	15:13
Air Blank	0.000	15:14
Control Test	0.081	15:14
Air Blank	0.000	15:14
Subject Sample # 1	0.058	15:15
Breath Volume	1.468	15:15
Air Blank	0.000	15:15
Air Blank	0.000	15:17
Subject Sample # 2	0.042	15:17
Breath Volume	3.781	15:17
Air Blank	0.000	15:18
Control Test	0.080	15:18
Air Blank	0.000	15:19
Diagnostics Check	OK	15:19

Data Download Date/Time: 06/26/2017 15:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005537

INSTRUMENT REGISTERED TO: DAYTONA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BISSA,JAMES,D
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: BISSA,JAMES
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 05/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14968
DRY GAS EXPIRY DATE: 08/06/2018
VIOLATION CODE: OTHER
ARREST TIME: 15:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:45

SUBJECT'S LAST NAME: TEST

DATE/TIME OF TEST: 06/23/2017 16:23:53

Diagnostics Check	OK	16:12	
Air Blank	0.000	16:12	
Control Test	0.080	16:12	
Air Blank	0.000	16:13	
Subject Sample	# 1 0.124	16:13	No .020 agreement
Breath Volume	1.570	16:13	
Air Blank	0.000	16:14	
Air Blank	0.000	16:16	
Subject Sample	# 2 0.090	16:16	No .020 agreement
Breath Volume	4.054	16:16	
Air Blank	0.000	16:17	
Air Blank	0.000	16:19	
Subject Sample	# 3	16:22	No Sample Provided
Breath Volume	0.000	16:22	
Air Blank	0.000	16:22	
Control Test	0.080	16:23	
Air Blank	0.000	16:23	
Diagnostics Check	OK	16:23	

Data Download Date/Time: 06/26/2017 15:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005537

INSTRUMENT REGISTERED TO: DAYTONA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAMPE,CORTLAND,B
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: THOMPSON,A
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 06/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14968
DRY GAS EXPIRY DATE: 08/06/2018
VIOLATION CODE: DUI
ARREST TIME: 01:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:39

SUBJECT'S LAST NAME: JINKINS

DATE/TIME OF TEST: 06/27/2017 03:06:34

Diagnostics Check	OK	02:59
Air Blank	0.000	03:00
Control Test	0.078	03:00
Air Blank	0.000	03:00
Subject Sample # 1	0.172	03:01
Breath Volume	2.625	03:01
Air Blank	0.000	03:02
Air Blank	0.000	03:04
Subject Sample # 2	0.164	03:04
Breath Volume	3.042	03:04
Air Blank	0.000	03:05
Control Test	0.078	03:05
Air Blank	0.000	03:06
Diagnostics Check	OK	03:06

Data Download Date/Time: 07/26/2017 15:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SLIMAK,M,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: SLIMAK,M
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 05/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: URBAN

DATE/TIME OF TEST: 06/01/2017 02:09:26

Diagnostics Check	OK	02:08	
Air Blank		02:08	RFI Detect
Air Blank	0.000	02:09	

Data Download Date/Time: 06/22/2017 11:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SLIMAK,M,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: SLIMAK,M
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 05/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: URBAN

DATE/TIME OF TEST: 06/01/2017 02:17:46

Diagnostics Check	OK	02:11
Air Blank	0.000	02:11
Control Test	0.079	02:11
Air Blank	0.000	02:12
Subject Sample # 1	0.120	02:13
Breath Volume	1.199	02:13
Air Blank	0.000	02:13
Air Blank	0.000	02:15
Subject Sample # 2	0.120	02:16
Breath Volume	1.335	02:16
Air Blank	0.000	02:16
Control Test	0.079	02:17
Air Blank	0.000	02:17
Diagnostics Check	OK	02:17

Data Download Date/Time: 06/22/2017 11:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAYES,BRITTNEY,B
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: HAYES,BRITTNEY
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 05/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 14:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:10

SUBJECT'S LAST NAME: MERIDA MERIDA

DATE/TIME OF TEST: 06/04/2017 16:07:40

Diagnostics Check	OK	15:53	
Air Blank	0.000	15:53	
Control Test	0.079	15:54	
Air Blank	0.000	15:54	
Subject Sample	# 1 0.147	15:57	Volume Not Met
Breath Volume	0.460	15:57	
Air Blank	0.000	15:58	
Air Blank	0.000	16:00	
Subject Sample	# 2 0.196	16:02	
Breath Volume	2.175	16:02	
Air Blank	0.000	16:03	
Air Blank	0.000	16:05	
Subject Sample	# 3 0.166	16:06	
Breath Volume	1.226	16:06	
Air Blank	0.000	16:06	
Control Test	0.069	16:07	Control Outside Tolerance
Air Blank	0.000	16:07	

Data Download Date/Time: 06/22/2017 11:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAYES,BRITTNEY,B
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: HAYES,BRITTNEY
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 05/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 14:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:10

SUBJECT'S LAST NAME: MERIDA MERIDA

DATE/TIME OF TEST: 06/04/2017 16:38:56

Diagnostics Check	OK	16:32
Air Blank	0.000	16:33
Control Test	0.078	16:33
Air Blank	0.000	16:33
Subject Sample # 1	0.186	16:34
Breath Volume	1.527	16:34
Air Blank	0.000	16:35
Air Blank	0.000	16:36
Subject Sample # 2	0.170	16:37
Breath Volume	1.484	16:37
Air Blank	0.000	16:38
Control Test	0.076	16:38
Air Blank	0.000	16:38
Diagnostics Check	OK	16:38

Data Download Date/Time: 06/22/2017 11:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MESA,JONATHAN,A
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: KEARNEY,A
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 05/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: CLARK

DATE/TIME OF TEST: 06/05/2017 03:03:44

Diagnostics Check	OK	03:00	
Air Blank	0.000	03:01	
Control Test	0.079	03:01	
Air Blank	0.000	03:02	
Subject Sample	# 1	03:02	Subject Test Refused
Breath Volume	0.000	03:02	
Air Blank	0.000	03:02	
Control Test	0.079	03:03	
Air Blank	0.000	03:03	
Diagnostics Check	OK	03:03	

Data Download Date/Time: 06/22/2017 11:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SLIMAK,M,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: SLIMAK,M
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 05/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 06/08/2017 02:18:07

Diagnostics Check	OK	02:17	
Air Blank		02:17	RFI Detect
Air Blank		02:18	RFI Detect

Data Download Date/Time: 06/22/2017 11:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SLIMAK,M,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: SLIMAK,M
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 05/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 06/08/2017 02:22:26

Diagnostics Check	OK	02:19	
Air Blank	0.000	02:20	
Control Test	0.079	02:20	
Air Blank	0.000	02:20	
Subject Sample	# 1	02:21	Subject Test Refused
Breath Volume	0.000	02:21	
Air Blank	0.000	02:21	
Control Test	0.079	02:21	
Air Blank	0.000	02:22	
Diagnostics Check	OK	02:22	

Data Download Date/Time: 06/22/2017 11:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SLIMAK,M,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: SLIMAK,M
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 05/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 05:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:05

SUBJECT'S LAST NAME: AGUILERA

DATE/TIME OF TEST: 06/09/2017 05:32:01

Diagnostics Check	OK	05:25
Air Blank	0.000	05:26
Control Test	0.079	05:26
Air Blank	0.000	05:27
Subject Sample # 1	0.122	05:27
Breath Volume	2.941	05:27
Air Blank	0.000	05:28
Air Blank	0.000	05:30
Subject Sample # 2	0.120	05:30
Breath Volume	2.984	05:30
Air Blank	0.000	05:31
Control Test	0.080	05:31
Air Blank	0.000	05:31
Diagnostics Check	OK	05:32

Data Download Date/Time: 06/22/2017 11:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 05/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: LANTIGUA

DATE/TIME OF TEST: 06/01/2017 01:01:42

Diagnostics Check	OK	00:55
Air Blank	0.000	00:56
Control Test	0.077	00:56
Air Blank	0.000	00:56
Subject Sample # 1	0.000	00:57
Breath Volume	2.753	00:57
Air Blank	0.000	00:57
Air Blank	0.000	00:59
Subject Sample # 2	0.000	01:00
Breath Volume	2.652	01:00
Air Blank	0.000	01:00
Control Test	0.077	01:01
Air Blank	0.000	01:01
Diagnostics Check	OK	01:01

Data Download Date/Time: 06/22/2017 11:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: CLOSIUS,RICHARD
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 05/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:11

SUBJECT'S LAST NAME: ALFARO SANCHEZ

DATE/TIME OF TEST: 06/01/2017 02:38:42

Diagnostics Check	OK	02:32
Air Blank	0.000	02:33
Control Test	0.077	02:33
Air Blank	0.000	02:33
Subject Sample # 1	0.204	02:34
Breath Volume	1.441	02:34
Air Blank	0.000	02:34
Air Blank	0.000	02:36
Subject Sample # 2	0.205	02:37
Breath Volume	2.378	02:37
Air Blank	0.000	02:37
Control Test	0.078	02:38
Air Blank	0.000	02:38
Diagnostics Check	OK	02:38

Data Download Date/Time: 06/22/2017 11:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: CLOSIUS,RICHARD
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 05/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: CRUZ-VAZQUEZ

DATE/TIME OF TEST: 06/02/2017 02:52:26

Diagnostics Check	OK	02:46
Air Blank	0.000	02:46
Control Test	0.077	02:47
Air Blank	0.000	02:47
Subject Sample # 1	0.057	02:48
Breath Volume	2.250	02:48
Air Blank	0.000	02:48
Air Blank	0.000	02:50
Subject Sample # 2	0.056	02:50
Breath Volume	3.214	02:50
Air Blank	0.000	02:51
Control Test	0.077	02:51
Air Blank	0.000	02:52
Diagnostics Check	OK	02:52

Data Download Date/Time: 06/22/2017 11:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 05/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:37

SUBJECT'S LAST NAME: TAVERAS

DATE/TIME OF TEST: 06/03/2017 04:03:53

Diagnostics Check	OK	04:01	
Air Blank	0.000	04:01	
Control Test	0.077	04:01	
Air Blank	0.000	04:02	
Subject Sample	# 1	04:02	Subject Test Refused
Breath Volume	0.000	04:02	
Air Blank	0.000	04:02	
Control Test	0.077	04:03	
Air Blank	0.000	04:03	
Diagnostics Check	OK	04:03	

Data Download Date/Time: 06/22/2017 11:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: CLOSIUS,RICHARD
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 05/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: FRANCEZA

DATE/TIME OF TEST: 06/04/2017 03:33:26

Diagnostics Check	OK	03:30	
Air Blank	0.000	03:31	
Control Test	0.077	03:31	
Air Blank	0.000	03:31	
Subject Sample	# 1	03:32	Subject Test Refused
Breath Volume	0.000	03:32	
Air Blank	0.000	03:32	
Control Test	0.076	03:32	
Air Blank	0.000	03:33	
Diagnostics Check	OK	03:33	

Data Download Date/Time: 06/22/2017 11:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: CLOSIUS,RICHARD
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 05/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:28

SUBJECT'S LAST NAME: PERFUMO

DATE/TIME OF TEST: 06/04/2017 04:54:08

Diagnostics Check	OK	04:51	
Air Blank	0.000	04:51	
Control Test	0.077	04:52	
Air Blank	0.000	04:52	
Subject Sample	# 1	04:52	Subject Test Refused
Breath Volume	0.000	04:52	
Air Blank	0.000	04:53	
Control Test	0.077	04:53	
Air Blank	0.000	04:54	
Diagnostics Check	OK	04:54	

Data Download Date/Time: 06/22/2017 11:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: CLOSIUS,RICHARD
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 05/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:11

SUBJECT'S LAST NAME: CASTILLO

DATE/TIME OF TEST: 06/10/2017 00:40:54

Diagnostics Check	OK	00:37	
Air Blank	0.000	00:38	
Control Test	0.077	00:38	
Air Blank	0.000	00:39	
Subject Sample	# 1	00:39	Subject Test Refused
Breath Volume	0.000	00:39	
Air Blank	0.000	00:39	
Control Test	0.077	00:40	
Air Blank	0.000	00:40	
Diagnostics Check	OK	00:40	

Data Download Date/Time: 06/22/2017 11:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 05/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: CASANOVA

DATE/TIME OF TEST: 06/10/2017 03:28:39

Diagnostics Check	OK	03:21
Air Blank	0.000	03:22
Control Test	0.077	03:22
Air Blank	0.000	03:23
Subject Sample # 1	0.135	03:23
Breath Volume	3.187	03:23
Air Blank	0.000	03:24
Air Blank	0.000	03:26
Subject Sample # 2	0.119	03:27
Breath Volume	1.175	03:27
Air Blank	0.000	03:27
Control Test	0.076	03:28
Air Blank	0.000	03:28
Diagnostics Check	OK	03:28

Data Download Date/Time: 06/22/2017 11:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: FRANCO,GREGORY
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 05/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:31

SUBJECT'S LAST NAME: RESTREPO

DATE/TIME OF TEST: 06/15/2017 03:57:20

Diagnostics Check	OK	03:54	
Air Blank	0.000	03:54	
Control Test	0.077	03:55	
Air Blank	0.000	03:55	
Subject Sample	# 1	03:55	Subject Test Refused
Breath Volume	0.000	03:55	
Air Blank	0.000	03:56	
Control Test	0.077	03:56	
Air Blank	0.000	03:57	
Diagnostics Check	OK	03:57	

Data Download Date/Time: 06/22/2017 11:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 05/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:22

SUBJECT'S LAST NAME: HERRERA

DATE/TIME OF TEST: 06/16/2017 03:11:11

Diagnostics Check	OK	03:02
Air Blank	0.000	03:03
Control Test	0.077	03:03
Air Blank	0.000	03:03
Subject Sample # 1	0.173	03:05
Breath Volume	2.496	03:05
Air Blank	0.000	03:06
Air Blank	0.000	03:08
Subject Sample # 2	0.163	03:09
Breath Volume	2.070	03:09
Air Blank	0.000	03:10
Control Test	0.075	03:10
Air Blank	0.000	03:11
Diagnostics Check	OK	03:11

Data Download Date/Time: 06/22/2017 11:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: FRANCO,GREGORY
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 05/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:26

SUBJECT'S LAST NAME: LONDONO

DATE/TIME OF TEST: 06/16/2017 04:49:48

Diagnostics Check	OK	04:46	
Air Blank	0.000	04:47	
Control Test	0.076	04:47	
Air Blank	0.000	04:48	
Subject Sample	# 1	04:48	Subject Test Refused
Breath Volume	0.000	04:48	
Air Blank	0.000	04:48	
Control Test	0.077	04:49	
Air Blank	0.000	04:49	
Diagnostics Check	OK	04:49	

Data Download Date/Time: 06/22/2017 11:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: CLOSIUS,RICHARD
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 05/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 05:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:16

SUBJECT'S LAST NAME: BLANCARTE

DATE/TIME OF TEST: 06/16/2017 06:43:53

Diagnostics Check	OK	06:36
Air Blank	0.000	06:37
Control Test	0.077	06:37
Air Blank	0.000	06:38
Subject Sample # 1	0.166	06:39
Breath Volume	2.187	06:39
Air Blank	0.000	06:39
Air Blank	0.000	06:41
Subject Sample # 2	0.154	06:42
Breath Volume	1.777	06:42
Air Blank	0.000	06:42
Control Test	0.076	06:43
Air Blank	0.000	06:43
Diagnostics Check	OK	06:43

Data Download Date/Time: 06/22/2017 11:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRANCO,GREGORY,R
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: FRANCO,GREGORY
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 05/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: MUNZON-TOMAS

DATE/TIME OF TEST: 06/17/2017 03:21:52

Diagnostics Check	OK	03:14
Air Blank	0.000	03:15
Control Test	0.077	03:15
Air Blank	0.000	03:16
Subject Sample # 1	0.128	03:17
Breath Volume	1.375	03:17
Air Blank	0.000	03:17
Air Blank	0.000	03:19
Subject Sample # 2	0.127	03:20
Breath Volume	2.425	03:20
Air Blank	0.000	03:20
Control Test	0.076	03:21
Air Blank	0.000	03:21
Diagnostics Check	OK	03:21

Data Download Date/Time: 06/22/2017 11:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: VARGAS,DAEL
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 05/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:19

SUBJECT'S LAST NAME: IBANEZ

DATE/TIME OF TEST: 06/17/2017 05:01:02

Diagnostics Check	OK	04:54
Air Blank	0.000	04:54
Control Test	0.076	04:55
Air Blank	0.000	04:55
Subject Sample # 1	0.161	04:56
Breath Volume	1.707	04:56
Air Blank	0.000	04:56
Air Blank	0.000	04:58
Subject Sample # 2	0.150	04:59
Breath Volume	2.117	04:59
Air Blank	0.000	05:00
Control Test	0.076	05:00
Air Blank	0.000	05:00
Diagnostics Check	OK	05:01

Data Download Date/Time: 06/22/2017 11:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZRAK,JOSEPH,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: AZRAK,JOSEPH
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 05/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: QUINONES

DATE/TIME OF TEST: 06/01/2017 01:02:01

Diagnostics Check	OK	00:58	
Air Blank	0.000	00:59	
Control Test	0.077	00:59	
Air Blank	0.000	00:59	
Subject Sample	# 1	01:00	Subject Test Refused
Breath Volume	0.000	01:00	
Air Blank	0.000	01:01	
Control Test	0.077	01:01	
Air Blank	0.000	01:01	
Diagnostics Check	OK	01:02	

Data Download Date/Time: 06/22/2017 14:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZRAK,JOSEPH,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: AZRAK,JOSEPH
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 05/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:39

SUBJECT'S LAST NAME: QUINONES

DATE/TIME OF TEST: 06/01/2017 02:07:20

Diagnostics Check	OK	01:59
Air Blank	0.000	01:59
Control Test	0.078	02:00
Air Blank	0.000	02:00
Subject Sample # 1	0.188	02:02
Breath Volume	2.343	02:02
Air Blank	0.000	02:03
Air Blank	0.000	02:05
Subject Sample # 2	0.186	02:05
Breath Volume	3.257	02:05
Air Blank	0.000	02:06
Control Test	0.076	02:06
Air Blank	0.000	02:07
Diagnostics Check	OK	02:07

Data Download Date/Time: 06/22/2017 14:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZRAK,JOSEPH,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: AZRAK,JOSEPH
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 05/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:53

SUBJECT'S LAST NAME: HUNT

DATE/TIME OF TEST: 06/04/2017 00:26:43

Diagnostics Check	OK	00:19
Air Blank	0.000	00:20
Control Test	0.078	00:20
Air Blank	0.000	00:21
Subject Sample # 1	0.284	00:22
Breath Volume	2.796	00:22
Air Blank	0.000	00:22
Air Blank	0.000	00:24
Subject Sample # 2	0.275	00:25
Breath Volume	2.605	00:25
Air Blank	0.000	00:25
Control Test	0.077	00:26
Air Blank	0.000	00:26
Diagnostics Check	OK	00:26

Data Download Date/Time: 06/22/2017 14:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZRAK,JOSEPH,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: AZRAK,JOSEPH
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 05/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: SANDOVAL

DATE/TIME OF TEST: 06/11/2017 04:24:08

Diagnostics Check	OK	04:20	
Air Blank	0.000	04:21	
Control Test	0.078	04:21	
Air Blank	0.000	04:22	
Subject Sample	# 1	04:22	Subject Test Refused
Breath Volume	0.000	04:22	
Air Blank	0.000	04:23	
Control Test	0.077	04:23	
Air Blank	0.000	04:24	
Diagnostics Check	OK	04:24	

Data Download Date/Time: 06/22/2017 14:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZRAK,JOSEPH,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: AZRAK,JOSEPH
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 05/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:18

SUBJECT'S LAST NAME: MORENO

DATE/TIME OF TEST: 06/22/2017 04:59:16

Diagnostics Check	OK	04:47	
Air Blank	0.000	04:47	
Control Test	0.077	04:47	
Air Blank	0.000	04:48	
Subject Sample	# 1 0.170	04:51	Volume Not Met
Breath Volume	0.085	04:51	
Air Blank	0.000	04:52	
Air Blank	0.000	04:53	
Subject Sample	# 2 0.181	04:54	
Breath Volume	2.085	04:54	
Air Blank	0.000	04:55	
Air Blank	0.000	04:57	
Subject Sample	# 3 0.186	04:57	
Breath Volume	1.296	04:57	
Air Blank	0.000	04:58	
Control Test	0.076	04:58	
Air Blank	0.000	04:59	
Diagnostics Check	OK	04:59	

Data Download Date/Time: 06/22/2017 14:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005655

INSTRUMENT REGISTERED TO: FORT MEYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOULD, LEONARD,
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: GOULD, LEONARD
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 05/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 15:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:10

SUBJECT'S LAST NAME: BIRCHER

DATE/TIME OF TEST: 06/01/2017 16:37:20

Diagnostics Check	OK	16:30
Air Blank	0.000	16:30
Control Test	0.079	16:31
Air Blank	0.000	16:31
Subject Sample # 1	0.196	16:32
Breath Volume	2.320	16:32
Air Blank	0.000	16:33
Air Blank	0.000	16:34
Subject Sample # 2	0.200	16:35
Breath Volume	1.906	16:35
Air Blank	0.000	16:36
Control Test	0.079	16:36
Air Blank	0.000	16:37
Diagnostics Check	OK	16:37

Data Download Date/Time: 06/02/2017 11:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005655

INSTRUMENT REGISTERED TO: FORT MEYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PINGRY,BRIAN,J
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: KOVACH,CHRISTOPHER
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 05/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 15:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:03

SUBJECT'S LAST NAME: CARROLL

DATE/TIME OF TEST: 06/01/2017 18:31:48

Diagnostics Check	OK	18:25
Air Blank	0.000	18:25
Control Test	0.079	18:25
Air Blank	0.000	18:26
Subject Sample # 1	0.191	18:26
Breath Volume	3.382	18:26
Air Blank	0.000	18:27
Air Blank	0.000	18:29
Subject Sample # 2	0.189	18:29
Breath Volume	3.585	18:29
Air Blank	0.000	18:30
Control Test	0.077	18:31
Air Blank	0.000	18:31
Diagnostics Check	OK	18:31

Data Download Date/Time: 06/02/2017 11:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005655

INSTRUMENT REGISTERED TO: FORT MEYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,JOSIAH,R
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: LOOS,S
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 05/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: GOOD

DATE/TIME OF TEST: 06/01/2017 23:29:46

Diagnostics Check	OK	23:26	
Air Blank	0.000	23:27	
Control Test	0.079	23:27	
Air Blank	0.000	23:28	
Subject Sample	# 1	23:28	Subject Test Refused
Breath Volume	0.000	23:28	
Air Blank	0.000	23:28	
Control Test	0.079	23:29	
Air Blank	0.000	23:29	
Diagnostics Check	OK	23:29	

Data Download Date/Time: 06/02/2017 11:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005655

INSTRUMENT REGISTERED TO: FORT MEYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALGER, MICHAEL, C
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: BROWN, E
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 05/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:56

SUBJECT'S LAST NAME: DOWELL

DATE/TIME OF TEST: 06/02/2017 03:33:38

Diagnostics Check	OK	03:26
Air Blank	0.000	03:27
Control Test	0.079	03:27
Air Blank	0.000	03:28
Subject Sample # 1	0.097	03:28
Breath Volume	2.277	03:28
Air Blank	0.000	03:29
Air Blank	0.000	03:31
Subject Sample # 2	0.096	03:31
Breath Volume	2.113	03:31
Air Blank	0.000	03:32
Control Test	0.078	03:32
Air Blank	0.000	03:33
Diagnostics Check	OK	03:33

Data Download Date/Time: 06/02/2017 11:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005655

INSTRUMENT REGISTERED TO: FORT MEYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUNCAN,N,
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: DUNCAN,N
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:26

SUBJECT'S LAST NAME: MCMULLEN

DATE/TIME OF TEST: 06/02/2017 23:56:10

Diagnostics Check	OK	23:49
Air Blank	0.000	23:50
Control Test	0.079	23:50
Air Blank	0.000	23:51
Subject Sample # 1	0.120	23:51
Breath Volume	3.207	23:51
Air Blank	0.000	23:52
Air Blank	0.000	23:53
Subject Sample # 2	0.114	23:54
Breath Volume	2.445	23:54
Air Blank	0.000	23:55
Control Test	0.078	23:55
Air Blank	0.000	23:56
Diagnostics Check	OK	23:56

Data Download Date/Time: 07/26/2017 12:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005655

INSTRUMENT REGISTERED TO: FORT MEYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,BILLY,S
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: ELLIS,BILLY
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 06/03/2017 02:28:32

Diagnostics Check	OK	02:21
Air Blank	0.000	02:21
Control Test	0.079	02:22
Air Blank	0.000	02:22
Subject Sample # 1	0.234	02:23
Breath Volume	1.328	02:23
Air Blank	0.000	02:24
Air Blank	0.000	02:26
Subject Sample # 2	0.233	02:26
Breath Volume	1.203	02:26
Air Blank	0.000	02:27
Control Test	0.076	02:27
Air Blank	0.000	02:28
Diagnostics Check	OK	02:28

Data Download Date/Time: 07/26/2017 12:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005655

INSTRUMENT REGISTERED TO: FORT MEYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLARK,JOSHUA,P
BREATH TEST OPERATOR AGENCY: Lee Port Authority
ARRESTING OFFICER: CLARK,JOSHUA
ARRESTING OFFICER AGENCY: Lee Port Authority
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: PEARL

DATE/TIME OF TEST: 06/03/2017 03:02:59

Diagnostics Check	OK	02:55
Air Blank	0.000	02:56
Control Test	0.078	02:56
Air Blank	0.000	02:57
Subject Sample # 1	0.119	02:58
Breath Volume	2.281	02:58
Air Blank	0.000	02:59
Air Blank	0.000	03:00
Subject Sample # 2	0.122	03:01
Breath Volume	1.750	03:01
Air Blank	0.000	03:02
Control Test	0.078	03:02
Air Blank	0.000	03:02
Diagnostics Check	OK	03:02

Data Download Date/Time: 07/26/2017 12:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005655

INSTRUMENT REGISTERED TO: FORT MEYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUNCAN,N,
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: DUNCAN,N
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:26

SUBJECT'S LAST NAME: HANDS

DATE/TIME OF TEST: 06/03/2017 03:56:47

Diagnostics Check	OK	03:50
Air Blank	0.000	03:50
Control Test	0.078	03:51
Air Blank	0.000	03:51
Subject Sample # 1	0.026	03:52
Breath Volume	1.957	03:52
Air Blank	0.000	03:52
Air Blank	0.000	03:54
Subject Sample # 2	0.025	03:55
Breath Volume	2.308	03:55
Air Blank	0.000	03:55
Control Test	0.078	03:56
Air Blank	0.000	03:56
Diagnostics Check	OK	03:56

Data Download Date/Time: 07/26/2017 12:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005655

INSTRUMENT REGISTERED TO: FORT MEYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUNCAN,N,
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: MCBRIDE,T
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: LIRANZO DE LEON

DATE/TIME OF TEST: 06/03/2017 04:06:45

Diagnostics Check	OK	04:03	
Air Blank	0.000	04:04	
Control Test	0.078	04:04	
Air Blank	0.000	04:04	
Subject Sample	# 1	04:05	Subject Test Refused
Breath Volume	0.000	04:05	
Air Blank	0.000	04:05	
Control Test	0.079	04:06	
Air Blank	0.000	04:06	
Diagnostics Check	OK	04:06	

Data Download Date/Time: 07/26/2017 12:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005655

INSTRUMENT REGISTERED TO: FORT MEYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,BILLY,S
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: PIERSON,CORY
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:23

SUBJECT'S LAST NAME: SHEPARD

DATE/TIME OF TEST: 06/03/2017 04:53:09

Diagnostics Check	OK	04:45
Air Blank	0.000	04:46
Control Test	0.078	04:46
Air Blank	0.000	04:46
Subject Sample # 1	0.011	04:48
Breath Volume	2.222	04:48
Air Blank	0.000	04:48
Air Blank	0.000	04:50
Subject Sample # 2	0.008	04:51
Breath Volume	2.156	04:51
Air Blank	0.000	04:52
Control Test	0.077	04:52
Air Blank	0.000	04:53
Diagnostics Check	OK	04:53

Data Download Date/Time: 07/26/2017 12:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005655

INSTRUMENT REGISTERED TO: FORT MEYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUNCAN,N,
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: DUNCAN,N
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: MOYER

DATE/TIME OF TEST: 06/04/2017 03:33:25

Diagnostics Check	OK	03:26
Air Blank	0.000	03:27
Control Test	0.080	03:27
Air Blank	0.000	03:28
Subject Sample # 1	0.113	03:28
Breath Volume	2.757	03:28
Air Blank	0.000	03:29
Air Blank	0.000	03:31
Subject Sample # 2	0.109	03:31
Breath Volume	2.992	03:31
Air Blank	0.000	03:32
Control Test	0.079	03:32
Air Blank	0.000	03:33
Diagnostics Check	OK	03:33

Data Download Date/Time: 07/26/2017 12:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005655

INSTRUMENT REGISTERED TO: FORT MEYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRASK,J,
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: TRASK,J
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 17:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:30

SUBJECT'S LAST NAME: COX

DATE/TIME OF TEST: 06/04/2017 17:59:38

Diagnostics Check	OK	17:52
Air Blank	0.000	17:52
Control Test	0.081	17:52
Air Blank	0.000	17:53
Subject Sample # 1	0.259	17:54
Breath Volume	2.937	17:54
Air Blank	0.000	17:55
Air Blank	0.000	17:56
Subject Sample # 2	0.255	17:57
Breath Volume	2.476	17:57
Air Blank	0.000	17:58
Control Test	0.077	17:58
Air Blank	0.000	17:59
Diagnostics Check	OK	17:59

Data Download Date/Time: 07/26/2017 12:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005655

INSTRUMENT REGISTERED TO: FORT MEYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUNCAN,N,
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: DUNCAN,N
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 18:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:55

SUBJECT'S LAST NAME: BLACK

DATE/TIME OF TEST: 06/04/2017 20:24:57

Diagnostics Check	OK	20:18
Air Blank	0.000	20:18
Control Test	0.079	20:18
Air Blank	0.000	20:19
Subject Sample # 1	0.235	20:20
Breath Volume	2.179	20:20
Air Blank	0.000	20:21
Air Blank	0.000	20:22
Subject Sample # 2	0.224	20:23
Breath Volume	1.925	20:23
Air Blank	0.000	20:23
Control Test	0.078	20:24
Air Blank	0.000	20:24
Diagnostics Check	OK	20:24

Data Download Date/Time: 07/26/2017 12:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005655

INSTRUMENT REGISTERED TO: FORT MEYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUNCAN,N,
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: DUNCAN,N
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:22

SUBJECT'S LAST NAME: RODRIGUEZ OLVERA

DATE/TIME OF TEST: 06/05/2017 02:51:56

Diagnostics Check	OK	02:44	
Air Blank	0.000	02:44	
Control Test	0.080	02:44	
Air Blank	0.000	02:45	
Subject Sample # 1	0.232	02:45	
Breath Volume	1.683	02:45	
Air Blank	0.000	02:46	
Air Blank	0.000	02:48	
Subject Sample # 2	0.200	02:50	
Breath Volume	1.511	02:50	
Air Blank		02:51	Purge Fail
Air Blank	0.000	02:51	

Data Download Date/Time: 07/26/2017 12:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005655

INSTRUMENT REGISTERED TO: FORT MEYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOMEZ,HENRRY,
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: BROWN,EARTHEN
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 17:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:22

SUBJECT'S LAST NAME: SELL

DATE/TIME OF TEST: 06/05/2017 19:02:27

Diagnostics Check	OK	18:58	
Air Blank	0.000	18:59	
Control Test	0.079	18:59	
Air Blank	0.000	19:00	
Subject Sample	# 1 0.245	19:01	
Breath Volume	1.886	19:01	
Air Blank		19:01	Purge Fail
Air Blank	0.000	19:02	

Data Download Date/Time: 07/26/2017 12:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005655

INSTRUMENT REGISTERED TO: FORT MEYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOMEZ,HENRRY,
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: BROWN,EARTHEN
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 17:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:22

SUBJECT'S LAST NAME: SELL

DATE/TIME OF TEST: 06/05/2017 19:05:28

Diagnostics Check	OK	19:04	
Air Blank	0.000	19:04	
Control Test	0.074	19:04	Control Outside Tolerance
Air Blank	0.000	19:05	

Data Download Date/Time: 07/26/2017 12:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005655

INSTRUMENT REGISTERED TO: FORT MEYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOMEZ,HENRRY,
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: BROWN,EARTHEN
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 17:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:22

SUBJECT'S LAST NAME: SELL

DATE/TIME OF TEST: 06/05/2017 19:23:18

Diagnostics Check	OK	19:17	
Air Blank	0.000	19:18	
Control Test	0.080	19:18	
Air Blank	0.000	19:18	
Subject Sample	# 1 0.166	19:22	Volume Not Met
Breath Volume	0.164	19:22	
Air Blank		19:22	Purge Fail
Air Blank	0.000	19:23	

Data Download Date/Time: 07/26/2017 12:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005655

INSTRUMENT REGISTERED TO: FORT MEYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUNCAN,N,
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: RALL,T
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 17:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:21

SUBJECT'S LAST NAME: ROMAN

DATE/TIME OF TEST: 06/06/2017 19:12:09

Diagnostics Check	OK	19:08	
Air Blank	0.000	19:09	
Control Test	0.080	19:09	
Air Blank	0.000	19:09	
Subject Sample	# 1	19:10	Interferent Detect
Breath Volume	1.414	19:10	
Air Blank	0.000	19:11	
Control Test	0.079	19:11	
Air Blank	0.000	19:12	
Diagnostics Check	OK	19:12	

Data Download Date/Time: 07/26/2017 12:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005655

INSTRUMENT REGISTERED TO: FORT MEYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUNCAN,N,
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: LINDHOLM,J
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 17:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:02

SUBJECT'S LAST NAME: JUAN JUAN

DATE/TIME OF TEST: 06/06/2017 20:33:11

Diagnostics Check	OK	20:26
Air Blank	0.000	20:26
Control Test	0.079	20:26
Air Blank	0.000	20:27
Subject Sample # 1	0.236	20:27
Breath Volume	1.585	20:27
Air Blank	0.000	20:28
Air Blank	0.000	20:30
Subject Sample # 2	0.232	20:31
Breath Volume	1.449	20:31
Air Blank	0.000	20:32
Control Test	0.078	20:32
Air Blank	0.000	20:33
Diagnostics Check	OK	20:33

Data Download Date/Time: 07/26/2017 12:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005655

INSTRUMENT REGISTERED TO: FORT MEYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOTELLO,A,
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: MEIXSELL,MATT
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: PADGETT

DATE/TIME OF TEST: 06/09/2017 05:45:26

Diagnostics Check	OK	05:38
Air Blank	0.000	05:38
Control Test	0.079	05:38
Air Blank	0.000	05:39
Subject Sample # 1	0.110	05:40
Breath Volume	3.144	05:40
Air Blank	0.000	05:41
Air Blank	0.000	05:43
Subject Sample # 2	0.103	05:43
Breath Volume	4.027	05:43
Air Blank	0.000	05:44
Control Test	0.075	05:44
Air Blank	0.000	05:45
Diagnostics Check	OK	05:45

Data Download Date/Time: 07/26/2017 12:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005655

INSTRUMENT REGISTERED TO: FORT MEYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALGER, MICHAEL, C
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: GENDRON, N
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: SANDOVAL

DATE/TIME OF TEST: 06/09/2017 23:14:33

Diagnostics Check	OK	23:08
Air Blank	0.000	23:08
Control Test	0.080	23:09
Air Blank	0.000	23:09
Subject Sample # 1	0.000	23:10
Breath Volume	1.949	23:10
Air Blank	0.000	23:10
Air Blank	0.000	23:12
Subject Sample # 2	0.000	23:13
Breath Volume	2.074	23:13
Air Blank	0.000	23:13
Control Test	0.078	23:13
Air Blank	0.000	23:14
Diagnostics Check	OK	23:14

Data Download Date/Time: 07/26/2017 12:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005655

INSTRUMENT REGISTERED TO: FORT MEYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,JOSIAH,R
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: SMITH,JOSIAH
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 18:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:05

SUBJECT'S LAST NAME: MATHES

DATE/TIME OF TEST: 06/15/2017 20:32:01

Diagnostics Check	OK	20:25
Air Blank	0.000	20:26
Control Test	0.080	20:26
Air Blank	0.000	20:26
Subject Sample # 1	0.000	20:27
Breath Volume	4.171	20:27
Air Blank	0.000	20:28
Air Blank	0.000	20:30
Subject Sample # 2	0.000	20:30
Breath Volume	3.890	20:30
Air Blank	0.000	20:31
Control Test	0.079	20:31
Air Blank	0.000	20:31
Diagnostics Check	OK	20:32

Data Download Date/Time: 07/26/2017 12:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005655

INSTRUMENT REGISTERED TO: FORT MEYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAUGHERTY,KENYATTA,O
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: DAUGHERTY,KENYATTA
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 06/16/2017 00:39:24

Diagnostics Check	OK	00:27	
Air Blank	0.000	00:27	
Control Test	0.079	00:28	
Air Blank	0.000	00:28	
Subject Sample	# 1 0.060	00:31	Volume Not Met
Breath Volume	0.984	00:31	
Air Blank	0.000	00:32	
Air Blank	0.000	00:34	
Subject Sample	# 2 0.061	00:37	Volume Not Met
Breath Volume	0.000	00:37	
Air Blank	0.000	00:38	
Control Test	0.078	00:38	
Air Blank	0.000	00:39	
Diagnostics Check	OK	00:39	

Data Download Date/Time: 07/26/2017 12:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005655

INSTRUMENT REGISTERED TO: FORT MEYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NAYLOR, MAURICE, E
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: NAYLOR, MAURICE
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 18:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:20

SUBJECT'S LAST NAME: ALVAREZ

DATE/TIME OF TEST: 06/20/2017 20:00:20

Diagnostics Check	OK	19:53
Air Blank	0.000	19:54
Control Test	0.080	19:54
Air Blank	0.000	19:55
Subject Sample # 1	0.163	19:55
Breath Volume	2.175	19:55
Air Blank	0.000	19:56
Air Blank	0.000	19:57
Subject Sample # 2	0.156	19:58
Breath Volume	1.843	19:58
Air Blank	0.000	19:59
Control Test	0.079	19:59
Air Blank	0.000	20:00
Diagnostics Check	OK	20:00

Data Download Date/Time: 07/26/2017 12:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005655

INSTRUMENT REGISTERED TO: FORT MEYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NAYLOR, MAURICE, E
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: BROWN, EARTHEN
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 17:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:42

SUBJECT'S LAST NAME: ACKERMAN

DATE/TIME OF TEST: 06/24/2017 19:12:25

Diagnostics Check	OK	19:05
Air Blank	0.000	19:06
Control Test	0.079	19:06
Air Blank	0.000	19:07
Subject Sample # 1	0.055	19:07
Breath Volume	2.906	19:07
Air Blank	0.000	19:08
Air Blank	0.000	19:10
Subject Sample # 2	0.051	19:10
Breath Volume	2.214	19:10
Air Blank	0.000	19:11
Control Test	0.080	19:11
Air Blank	0.000	19:12
Diagnostics Check	OK	19:12

Data Download Date/Time: 07/26/2017 12:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005655

INSTRUMENT REGISTERED TO: FORT MEYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NAYLOR, MAURICE, E
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: BROWN, EARTHEN
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:02

SUBJECT'S LAST NAME: SAEZ

DATE/TIME OF TEST: 06/25/2017 00:23:08

Diagnostics Check	OK	00:16
Air Blank	0.000	00:17
Control Test	0.080	00:17
Air Blank	0.000	00:17
Subject Sample # 1	0.123	00:18
Breath Volume	3.496	00:18
Air Blank	0.000	00:19
Air Blank	0.000	00:20
Subject Sample # 2	0.121	00:21
Breath Volume	2.929	00:21
Air Blank	0.000	00:22
Control Test	0.079	00:22
Air Blank	0.000	00:23
Diagnostics Check	OK	00:23

Data Download Date/Time: 07/26/2017 12:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005655

INSTRUMENT REGISTERED TO: FORT MEYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHAEFER, KENDALL, A
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: REYNOLDS, C
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 18:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:30

SUBJECT'S LAST NAME: DEVORE

DATE/TIME OF TEST: 06/26/2017 20:41:18

Diagnostics Check	OK	20:34
Air Blank	0.000	20:34
Control Test	0.080	20:35
Air Blank	0.000	20:35
Subject Sample # 1	0.160	20:36
Breath Volume	2.824	20:36
Air Blank	0.000	20:37
Air Blank	0.000	20:38
Subject Sample # 2	0.164	20:39
Breath Volume	2.156	20:39
Air Blank	0.000	20:40
Control Test	0.078	20:40
Air Blank	0.000	20:41
Diagnostics Check	OK	20:41

Data Download Date/Time: 07/26/2017 12:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005655

INSTRUMENT REGISTERED TO: FORT MEYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHAEFER, KENDALL, A
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: WITT, KELLY
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:53

SUBJECT'S LAST NAME: NEWLAND II

DATE/TIME OF TEST: 06/26/2017 23:34:05

Diagnostics Check	OK	23:27
Air Blank	0.000	23:28
Control Test	0.079	23:28
Air Blank	0.000	23:29
Subject Sample # 1	0.000	23:29
Breath Volume	2.156	23:29
Air Blank	0.000	23:30
Air Blank	0.000	23:32
Subject Sample # 2	0.000	23:32
Breath Volume	2.253	23:32
Air Blank	0.000	23:33
Control Test	0.080	23:33
Air Blank	0.000	23:33
Diagnostics Check	OK	23:34

Data Download Date/Time: 07/26/2017 12:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005655

INSTRUMENT REGISTERED TO: FORT MEYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHAEFER, KENDALL, A
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: HERNANDEZ, YILKANIA
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:27

SUBJECT'S LAST NAME: RAMBISOON

DATE/TIME OF TEST: 06/28/2017 04:56:08

Diagnostics Check	OK	04:49	
Air Blank	0.000	04:49	
Control Test	0.079	04:50	
Air Blank	0.000	04:50	
Subject Sample # 1	0.254	04:51	
Breath Volume	2.253	04:51	
Air Blank	0.000	04:52	
Air Blank	0.000	04:53	
Subject Sample # 2	0.256	04:54	
Breath Volume	2.316	04:54	
Air Blank	0.000	04:55	
Control Test	0.070	04:55	Control Outside Tolerance
Air Blank	0.000	04:56	

Data Download Date/Time: 07/26/2017 12:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005655

INSTRUMENT REGISTERED TO: FORT MEYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENEMAN,LESA,
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: BRENEMAN,LESA
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:37

SUBJECT'S LAST NAME: MOTLEY

DATE/TIME OF TEST: 06/30/2017 22:01:34

Diagnostics Check	OK	21:58	
Air Blank	0.000	21:59	
Control Test	0.080	21:59	
Air Blank	0.000	22:00	
Subject Sample	# 1	22:00	Subject Test Refused
Breath Volume	0.000	22:00	
Air Blank	0.000	22:00	
Control Test	0.080	22:00	
Air Blank	0.000	22:01	
Diagnostics Check	OK	22:01	

Data Download Date/Time: 07/26/2017 12:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005935

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LYONS,D,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: CAMPOS,M
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 05/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 03:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:59

SUBJECT'S LAST NAME: SPRINGHAM

DATE/TIME OF TEST: 06/01/2017 04:25:41

Diagnostics Check	OK	04:19
Air Blank	0.000	04:19
Control Test	0.079	04:20
Air Blank	0.000	04:20
Subject Sample # 1	0.222	04:21
Breath Volume	3.316	04:21
Air Blank	0.000	04:21
Air Blank	0.000	04:23
Subject Sample # 2	0.228	04:24
Breath Volume	2.429	04:24
Air Blank	0.000	04:24
Control Test	0.079	04:25
Air Blank	0.000	04:25
Diagnostics Check	OK	04:25

Data Download Date/Time: 06/27/2017 19:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005935

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LYONS,D,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: LYONS,D
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 05/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 04:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: GONZALEZ PEDROGO

DATE/TIME OF TEST: 06/10/2017 04:47:18

Diagnostics Check	OK	04:41	
Air Blank	0.000	04:41	
Control Test	0.079	04:42	
Air Blank	0.000	04:42	
Subject Sample # 1	0.167	04:43	
Breath Volume	1.488	04:43	
Air Blank	0.000	04:43	
Air Blank	0.000	04:45	
		04:45	Improper Sample
Air Blank	0.000	04:46	
Control Test	0.078	04:46	
Air Blank	0.000	04:47	
Diagnostics Check	OK	04:47	

Data Download Date/Time: 06/27/2017 19:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005935

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LYONS],D,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: LYONS,D
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 05/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 04:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: GONZALEZ PEDROGO

DATE/TIME OF TEST: 06/10/2017 04:56:31

Diagnostics Check	OK	04:49
Air Blank	0.000	04:49
Control Test	0.079	04:50
Air Blank	0.000	04:50
Subject Sample # 1	0.166	04:51
Breath Volume	1.160	04:51
Air Blank	0.000	04:51
Air Blank	0.000	04:53
Subject Sample # 2	0.153	04:54
Breath Volume	1.320	04:54
Air Blank	0.000	04:55
Control Test	0.078	04:55
Air Blank	0.000	04:56
Diagnostics Check	OK	04:56

Data Download Date/Time: 06/27/2017 19:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005935

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LYONS,D,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: LYONS,D
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 05/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 06:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:30

SUBJECT'S LAST NAME: GRAVES

DATE/TIME OF TEST: 06/19/2017 06:58:28

Diagnostics Check	OK	06:50
Air Blank	0.000	06:50
Control Test	0.080	06:51
Air Blank	0.000	06:51
Subject Sample # 1	0.141	06:52
Breath Volume	1.648	06:52
Air Blank	0.000	06:53
Air Blank	0.000	06:55
Subject Sample # 2	0.148	06:56
Breath Volume	1.214	06:56
Air Blank	0.000	06:57
Control Test	0.079	06:57
Air Blank	0.000	06:58
Diagnostics Check	OK	06:58

Data Download Date/Time: 06/27/2017 19:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005935

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LYONS,D,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: LYONS,D
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 05/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 04:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:55

SUBJECT'S LAST NAME: JAISINGH

DATE/TIME OF TEST: 06/21/2017 05:24:32

Diagnostics Check	OK	05:15
Air Blank	0.000	05:15
Control Test	0.079	05:16
Air Blank	0.000	05:16
Subject Sample # 1	0.179	05:18
Breath Volume	1.125	05:18
Air Blank	0.000	05:18
Air Blank	0.000	05:20
Subject Sample # 2	0.174	05:22
Breath Volume	1.351	05:22
Air Blank	0.000	05:23
Control Test	0.077	05:23
Air Blank	0.000	05:24
Diagnostics Check	OK	05:24

Data Download Date/Time: 06/27/2017 19:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005935

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LYONS,D,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: LYONS,D
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 05/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 00:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:41

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 06/24/2017 01:16:09

Diagnostics Check	OK	01:08
Air Blank	0.000	01:08
Control Test	0.078	01:08
Air Blank	0.000	01:09
Subject Sample # 1	0.139	01:10
Breath Volume	1.511	01:10
Air Blank	0.000	01:11
Air Blank	0.000	01:13
Subject Sample # 2	0.125	01:14
Breath Volume	1.781	01:14
Air Blank	0.000	01:15
Control Test	0.078	01:15
Air Blank	0.000	01:16
Diagnostics Check	OK	01:16

Data Download Date/Time: 06/27/2017 19:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005935

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LYONS,D,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: LYONS,D
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 05/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 22:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:39

SUBJECT'S LAST NAME: OZA

DATE/TIME OF TEST: 06/24/2017 22:25:44

Diagnostics Check	OK	22:13	
Air Blank	0.000	22:13	
Control Test	0.079	22:14	
Air Blank	0.000	22:14	
Subject Sample	# 1 0.263	22:16	
Breath Volume	1.574	22:16	
Air Blank	0.000	22:17	
Air Blank	0.000	22:19	
Subject Sample	# 2 0.285	22:20	
Breath Volume	1.957	22:20	
Air Blank	0.000	22:21	
Air Blank	0.000	22:23	
Subject Sample	# 3 0.269	22:23	
Breath Volume	1.300	22:23	
Air Blank	0.000	22:24	
Control Test	0.077	22:24	
Air Blank	0.000	22:25	
Diagnostics Check	Fail	22:25	Voltage/Current Test
Air Blank	0.000	22:25	

Data Download Date/Time: 06/27/2017 19:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005935

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LYONS,D,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: LYONS,D
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 05/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 03:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: ROSERO

DATE/TIME OF TEST: 06/25/2017 03:21:41

Diagnostics Check	OK	03:15
Air Blank	0.000	03:15
Control Test	0.079	03:16
Air Blank	0.000	03:16
Subject Sample # 1	0.164	03:17
Breath Volume	1.500	03:17
Air Blank	0.000	03:17
Air Blank	0.000	03:19
Subject Sample # 2	0.174	03:20
Breath Volume	1.953	03:20
Air Blank	0.000	03:20
Control Test	0.079	03:21
Air Blank	0.000	03:21
Diagnostics Check	OK	03:21

Data Download Date/Time: 06/27/2017 19:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005936

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PUFAHL,G,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: PUFAHL,G
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 05/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:27

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 06/02/2017 01:00:15

Diagnostics Check	OK	00:47	
Air Blank	0.000	00:47	
Control Test	0.080	00:48	
Air Blank	0.000	00:48	
Subject Sample	# 1 0.238	00:51	Volume Not Met
Breath Volume	0.082	00:51	
Air Blank	0.000	00:52	
Air Blank	0.000	00:54	
Subject Sample	# 2 0.260	00:54	
Breath Volume	3.402	00:54	
Air Blank	0.000	00:55	
Air Blank	0.000	00:57	
Subject Sample	# 3 0.256	00:58	
Breath Volume	2.066	00:58	
Air Blank	0.000	00:59	
Control Test	0.078	00:59	
Air Blank	0.000	01:00	
Diagnostics Check	OK	01:00	

Data Download Date/Time: 06/28/2017 09:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALMER,SAMANTHA,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SCHNEIDER,JOSHUA
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 05/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:24

SUBJECT'S LAST NAME: DE SCENZA

DATE/TIME OF TEST: 06/01/2017 02:59:16

Diagnostics Check	OK	02:51
Air Blank	0.000	02:52
Control Test	0.081	02:52
Air Blank	0.000	02:53
Subject Sample # 1	0.169	02:54
Breath Volume	2.976	02:54
Air Blank	0.000	02:55
Air Blank	0.000	02:57
Subject Sample # 2	0.169	02:57
Breath Volume	1.945	02:57
Air Blank	0.000	02:58
Control Test	0.080	02:58
Air Blank	0.000	02:59
Diagnostics Check	OK	02:59

Data Download Date/Time: 06/02/2017 16:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALMER,SAMANTHA,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ESCARAN,LESTER
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 05/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:47

SUBJECT'S LAST NAME: LE

DATE/TIME OF TEST: 06/01/2017 04:19:48

Diagnostics Check	OK	04:12
Air Blank	0.000	04:13
Control Test	0.082	04:13
Air Blank	0.000	04:13
Subject Sample # 1	0.235	04:14
Breath Volume	2.277	04:14
Air Blank	0.000	04:15
Air Blank	0.000	04:17
Subject Sample # 2	0.235	04:18
Breath Volume	1.878	04:18
Air Blank	0.000	04:18
Control Test	0.080	04:19
Air Blank	0.000	04:19
Diagnostics Check	OK	04:19

Data Download Date/Time: 06/02/2017 16:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: HANDZIK,BIRGIT
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 05/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 15:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:15

SUBJECT'S LAST NAME: RICE

DATE/TIME OF TEST: 06/01/2017 17:27:38

Diagnostics Check	OK	17:24	
Air Blank	0.000	17:25	
Control Test	0.081	17:25	
Air Blank	0.000	17:25	
Subject Sample	# 1	17:26	Subject Test Refused
Breath Volume	0.000	17:26	
Air Blank	0.000	17:26	
Control Test	0.081	17:27	
Air Blank	0.000	17:27	
Diagnostics Check	OK	17:27	

Data Download Date/Time: 06/02/2017 16:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SCHNEIDER,JOSHUA
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 05/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:24

SUBJECT'S LAST NAME: MULET

DATE/TIME OF TEST: 06/01/2017 22:57:21

Diagnostics Check	OK	22:51
Air Blank	0.000	22:51
Control Test	0.081	22:51
Air Blank	0.000	22:52
Subject Sample # 1	0.000	22:53
Breath Volume	2.468	22:53
Air Blank	0.000	22:53
Air Blank	0.000	22:55
Subject Sample # 2	0.000	22:55
Breath Volume	1.636	22:55
Air Blank	0.000	22:56
Control Test	0.082	22:56
Air Blank	0.000	22:57
Diagnostics Check	OK	22:57

Data Download Date/Time: 06/02/2017 16:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: POINTU,PATRICK
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 05/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: ABRAMS

DATE/TIME OF TEST: 06/02/2017 02:33:29

Diagnostics Check	OK	02:26
Air Blank	0.000	02:27
Control Test	0.081	02:27
Air Blank	0.000	02:27
Subject Sample # 1	0.168	02:28
Breath Volume	2.386	02:28
Air Blank	0.000	02:29
Air Blank	0.000	02:31
Subject Sample # 2	0.160	02:31
Breath Volume	1.113	02:31
Air Blank	0.000	02:32
Control Test	0.080	02:32
Air Blank	0.000	02:33
Diagnostics Check	OK	02:33

Data Download Date/Time: 06/02/2017 16:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARI,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ESCARAN,LESTER
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 05/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:40

SUBJECT'S LAST NAME: VILLALTA

DATE/TIME OF TEST: 06/02/2017 05:21:42

Diagnostics Check	OK	05:05	
Air Blank	0.000	05:06	
Control Test	0.081	05:06	
Air Blank	0.000	05:07	
Subject Sample	# 1 0.153	05:08	
Breath Volume	1.152	05:08	
Air Blank	0.000	05:09	
Air Blank	0.000	05:11	
Subject Sample	# 2 0.139	05:14	Volume Not Met
Breath Volume	0.453	05:14	
Air Blank	0.000	05:15	
Air Blank	0.000	05:16	
Subject Sample	# 3 0.126	05:19	Volume Not Met
Breath Volume	0.289	05:19	
Air Blank	0.000	05:20	
Control Test	0.078	05:21	
Air Blank	0.000	05:21	
Diagnostics Check	OK	05:21	

Data Download Date/Time: 06/02/2017 16:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARI,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ESCARAN,LESTER
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 05/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:40

SUBJECT'S LAST NAME: VILLALTA

DATE/TIME OF TEST: 06/02/2017 05:37:40

Diagnostics Check	OK	05:22	
Air Blank	0.000	05:23	
Control Test	0.078	05:23	
Air Blank	0.000	05:24	
Subject Sample	# 1 0.141	05:24	
Breath Volume	1.597	05:24	
Air Blank	0.000	05:25	
Air Blank	0.000	05:27	
Subject Sample	# 2 0.140	05:30	Volume Not Met
Breath Volume	0.488	05:30	
Air Blank	0.000	05:31	
Air Blank	0.000	05:32	
Subject Sample	# 3 0.133	05:35	Volume Not Met
Breath Volume	0.367	05:35	
Air Blank	0.000	05:36	
Control Test	0.077	05:37	
Air Blank	0.000	05:37	
Diagnostics Check	OK	05:37	

Data Download Date/Time: 06/02/2017 16:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006085

INSTRUMENT REGISTERED TO: US COAST GUARD MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON, RONALD,
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: WASHINGTON, RONALD
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 05/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1826916
DRY GAS EXPIRY DATE: 03/09/2018
VIOLATION CODE: BUI
ARREST TIME: 19:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:05

SUBJECT'S LAST NAME: DE LA OSA

DATE/TIME OF TEST: 06/11/2017 19:40:08

Diagnostics Check	OK	19:35	
Air Blank	0.000	19:36	
Control Test	0.079	19:36	
Air Blank	0.000	19:37	
Subject Sample	# 1	19:38	Subject Test Refused
Breath Volume	0.000	19:38	
Air Blank	0.000	19:39	
Control Test	0.080	19:39	
Air Blank	0.000	19:40	
Diagnostics Check	OK	19:40	

Data Download Date/Time: 06/29/2017 13:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006085

INSTRUMENT REGISTERED TO: US COAST GUARD MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIBLE,TIMOTHY,J
BREATH TEST OPERATOR AGENCY: USCG Miami
ARRESTING OFFICER: SARMIENTO,MIKE
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 06/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1826916
DRY GAS EXPIRY DATE: 03/09/2018
VIOLATION CODE: BUI
ARREST TIME: 18:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:40

SUBJECT'S LAST NAME: LLOVIO

DATE/TIME OF TEST: 06/30/2017 19:34:49

Diagnostics Check	OK	19:28
Air Blank	0.000	19:29
Control Test	0.078	19:29
Air Blank	0.000	19:29
Subject Sample # 1	0.081	19:30
Breath Volume	3.058	19:30
Air Blank	0.000	19:31
Air Blank	0.000	19:32
Subject Sample # 2	0.082	19:33
Breath Volume	3.238	19:33
Air Blank	0.000	19:33
Control Test	0.078	19:34
Air Blank	0.000	19:34
Diagnostics Check	OK	19:34

Data Download Date/Time: 07/26/2017 11:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELLAS,DILLON,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: BELLAS,DILLON
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 06/16/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06616080A2
DRY GAS EXPIRY DATE: 04/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:52

SUBJECT'S LAST NAME: TABARES

DATE/TIME OF TEST: 06/17/2017 22:23:46

Diagnostics Check	OK	22:17
Air Blank	0.000	22:17
Control Test	0.081	22:17
Air Blank	0.000	22:18
Subject Sample # 1	0.134	22:19
Breath Volume	1.980	22:19
Air Blank	0.000	22:19
Air Blank	0.000	22:21
Subject Sample # 2	0.121	22:22
Breath Volume	1.968	22:22
Air Blank	0.000	22:22
Control Test	0.080	22:23
Air Blank	0.000	22:23
Diagnostics Check	OK	22:23

Data Download Date/Time: 07/24/2017 13:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELLAS,DILLON,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: BELLAS,DILLON
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 06/16/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06616080A2
DRY GAS EXPIRY DATE: 04/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:50

SUBJECT'S LAST NAME: HARDISTY

DATE/TIME OF TEST: 06/18/2017 20:17:37

Diagnostics Check	OK	20:11
Air Blank	0.000	20:11
Control Test	0.080	20:11
Air Blank	0.000	20:12
Subject Sample # 1	0.060	20:13
Breath Volume	2.828	20:13
Air Blank	0.000	20:13
Air Blank	0.000	20:15
Subject Sample # 2	0.058	20:16
Breath Volume	3.281	20:16
Air Blank	0.000	20:16
Control Test	0.080	20:17
Air Blank	0.000	20:17
Diagnostics Check	OK	20:17

Data Download Date/Time: 07/24/2017 13:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAY,ERICK,J
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: DAY,ERICK
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 06/16/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06616080A2
DRY GAS EXPIRY DATE: 04/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:04

SUBJECT'S LAST NAME: RICHTERS

DATE/TIME OF TEST: 06/23/2017 00:48:12

Diagnostics Check	OK	00:41
Air Blank	0.000	00:42
Control Test	0.081	00:42
Air Blank	0.000	00:43
Subject Sample # 1	0.212	00:43
Breath Volume	2.667	00:43
Air Blank	0.000	00:44
Air Blank	0.000	00:46
Subject Sample # 2	0.220	00:46
Breath Volume	2.570	00:46
Air Blank	0.000	00:47
Control Test	0.079	00:47
Air Blank	0.000	00:48
Diagnostics Check	OK	00:48

Data Download Date/Time: 07/24/2017 13:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELLAS,DILLON,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: BELLAS,DILLON
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 06/16/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06616080A2
DRY GAS EXPIRY DATE: 04/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: MENDOZA

DATE/TIME OF TEST: 06/25/2017 20:28:11

Diagnostics Check	OK	20:21
Air Blank	0.000	20:22
Control Test	0.081	20:22
Air Blank	0.000	20:23
Subject Sample # 1	0.273	20:23
Breath Volume	2.589	20:23
Air Blank	0.000	20:24
Air Blank	0.000	20:26
Subject Sample # 2	0.272	20:26
Breath Volume	2.253	20:26
Air Blank	0.000	20:27
Control Test	0.080	20:27
Air Blank	0.000	20:28
Diagnostics Check	OK	20:28

Data Download Date/Time: 07/24/2017 13:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANGERVIL,CLAUDY,P
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: ANGERVIL,CLAUDY
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 06/16/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06616080A2
DRY GAS EXPIRY DATE: 04/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:39

SUBJECT'S LAST NAME: JAMES

DATE/TIME OF TEST: 06/27/2017 00:02:08

Diagnostics Check	OK	23:59	
Air Blank	0.000	23:59	
Control Test	0.081	00:00	
Air Blank	0.000	00:00	
Subject Sample	# 1	00:00	Subject Test Refused
Breath Volume	0.000	00:00	
Air Blank	0.000	00:01	
Control Test	0.081	00:01	
Air Blank	0.000	00:02	
Diagnostics Check	OK	00:02	

Data Download Date/Time: 07/24/2017 13:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SERRA,CORP,
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: SERRA,CORP
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 06/16/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06616080A2
DRY GAS EXPIRY DATE: 04/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:15

SUBJECT'S LAST NAME: MUSZICK

DATE/TIME OF TEST: 06/28/2017 20:21:52

Diagnostics Check	OK	20:15
Air Blank	0.000	20:15
Control Test	0.081	20:16
Air Blank	0.000	20:16
Subject Sample # 1	0.285	20:17
Breath Volume	2.476	20:17
Air Blank	0.000	20:17
Air Blank	0.000	20:19
Subject Sample # 2	0.294	20:20
Breath Volume	2.355	20:20
Air Blank	0.000	20:20
Control Test	0.080	20:21
Air Blank	0.000	20:21
Diagnostics Check	OK	20:21

Data Download Date/Time: 07/24/2017 13:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POLIZZI,CHRISTINE,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: WISE,EUGENE
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 06/16/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06616080A2
DRY GAS EXPIRY DATE: 04/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: WALKER

DATE/TIME OF TEST: 06/28/2017 23:07:05

Diagnostics Check	OK	23:01
Air Blank	0.000	23:01
Control Test	0.082	23:01
Air Blank	0.000	23:02
Subject Sample # 1	0.000	23:02
Breath Volume	2.562	23:02
Air Blank	0.000	23:03
Air Blank	0.000	23:05
Subject Sample # 2	0.000	23:05
Breath Volume	1.890	23:05
Air Blank	0.000	23:06
Control Test	0.082	23:06
Air Blank	0.000	23:07
Diagnostics Check	OK	23:07

Data Download Date/Time: 07/24/2017 13:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PECCI,Z,A
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: PECCI,Z
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 06/16/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06616080A2
DRY GAS EXPIRY DATE: 04/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:33

SUBJECT'S LAST NAME: SNYDER

DATE/TIME OF TEST: 06/30/2017 01:00:14

Diagnostics Check	OK	00:53
Air Blank	0.000	00:54
Control Test	0.081	00:54
Air Blank	0.000	00:54
Subject Sample # 1	0.172	00:55
Breath Volume	1.285	00:55
Air Blank	0.000	00:56
Air Blank	0.000	00:58
Subject Sample # 2	0.170	00:58
Breath Volume	1.632	00:58
Air Blank	0.000	00:59
Control Test	0.080	00:59
Air Blank	0.000	01:00
Diagnostics Check	OK	01:00

Data Download Date/Time: 07/24/2017 13:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POLIZZI,CHRISTINE,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: POLIZZI,CHRISTINE
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 06/16/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06616080A2
DRY GAS EXPIRY DATE: 04/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: BUFKIN

DATE/TIME OF TEST: 06/30/2017 01:22:25

Diagnostics Check	OK	01:15
Air Blank	0.000	01:16
Control Test	0.081	01:16
Air Blank	0.000	01:16
Subject Sample # 1	0.114	01:17
Breath Volume	1.351	01:17
Air Blank	0.000	01:18
Air Blank	0.000	01:19
Subject Sample # 2	0.114	01:20
Breath Volume	1.312	01:20
Air Blank	0.000	01:21
Control Test	0.080	01:21
Air Blank	0.000	01:22
Diagnostics Check	OK	01:22

Data Download Date/Time: 07/24/2017 13:44