

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000868

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALVAREZ,ORLANDO,
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: ALVAREZ,ORLANDO
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 06/14/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 562838
DRY GAS EXPIRY DATE: 09/06/2019
VIOLATION CODE: DUI
ARREST TIME: 00:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: KRINGETA

DATE/TIME OF TEST: 06/17/2017 01:14:27

Diagnostics Check	OK	01:10	
Air Blank	0.000	01:11	
Control Test	0.080	01:11	
Air Blank	0.000	01:12	
Subject Sample	# 1	01:13	Subject Test Refused
Breath Volume	0.000	01:13	
Air Blank	0.000	01:13	
Control Test	0.081	01:13	
Air Blank	0.000	01:14	
Diagnostics Check	OK	01:14	

Data Download Date/Time: 07/21/2017 16:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000868

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOPP,SETH,M
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: HOPP,SETH
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 06/14/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 562838
DRY GAS EXPIRY DATE: 09/06/2019
VIOLATION CODE: DUI
ARREST TIME: 23:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: STEENSTRA

DATE/TIME OF TEST: 06/18/2017 00:21:24

Diagnostics Check	OK	00:18	
Air Blank	0.000	00:18	
Control Test	0.080	00:19	
Air Blank	0.000	00:19	
Subject Sample	# 1	00:20	Subject Test Refused
Breath Volume	0.000	00:20	
Air Blank	0.000	00:20	
Control Test	0.081	00:20	
Air Blank	0.000	00:21	
Diagnostics Check	OK	00:21	

Data Download Date/Time: 07/21/2017 16:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000868

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LONES,SHAUN,G
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: LONES,SHAUN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 06/14/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 562838
DRY GAS EXPIRY DATE: 09/06/2019
VIOLATION CODE: DUI
ARREST TIME: 23:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:14

SUBJECT'S LAST NAME: LASSETTER

DATE/TIME OF TEST: 06/24/2017 00:37:00

Diagnostics Check	OK	00:34	
Air Blank	0.000	00:34	
Control Test	0.081	00:34	
Air Blank	0.000	00:35	
Subject Sample	# 1	00:35	Subject Test Refused
Breath Volume	0.000	00:35	
Air Blank	0.000	00:36	
Control Test	0.081	00:36	
Air Blank	0.000	00:36	
Diagnostics Check	OK	00:37	

Data Download Date/Time: 07/21/2017 16:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000868

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN-REYES, JUAN, C
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: DANIELS, WILLIAM
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 06/14/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 562838
DRY GAS EXPIRY DATE: 09/06/2019
VIOLATION CODE: DUI
ARREST TIME: 19:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:36

SUBJECT'S LAST NAME: PARK

DATE/TIME OF TEST: 06/29/2017 20:02:36

Diagnostics Check	OK	19:56
Air Blank	0.000	19:56
Control Test	0.081	19:57
Air Blank	0.000	19:57
Subject Sample # 1	0.149	19:58
Breath Volume	1.199	19:58
Air Blank	0.000	19:58
Air Blank	0.000	20:00
Subject Sample # 2	0.146	20:01
Breath Volume	1.589	20:01
Air Blank	0.000	20:01
Control Test	0.080	20:02
Air Blank	0.000	20:02
Diagnostics Check	OK	20:02

Data Download Date/Time: 07/21/2017 16:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,REINIER,
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: PEREZ,REINIER
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 05/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14968
DRY GAS EXPIRY DATE: 08/06/2018
VIOLATION CODE: DUI
ARREST TIME: 21:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: WELLS

DATE/TIME OF TEST: 06/01/2017 22:37:12

Diagnostics Check	OK	22:34	
Air Blank	0.000	22:34	
Control Test	0.079	22:34	
Air Blank	0.000	22:35	
Subject Sample	# 1	22:35	Subject Test Refused
Breath Volume	0.000	22:35	
Air Blank	0.000	22:36	
Control Test	0.081	22:36	
Air Blank	0.000	22:37	
Diagnostics Check	OK	22:37	

Data Download Date/Time: 06/29/2017 12:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,ALEX,
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: RODRIGUEZ,ALEX
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 05/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14968
DRY GAS EXPIRY DATE: 08/06/2018
VIOLATION CODE: DUI
ARREST TIME: 23:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:29

SUBJECT'S LAST NAME: HIGGINS

DATE/TIME OF TEST: 06/02/2017 00:56:22

Diagnostics Check	OK	00:53	
Air Blank	0.000	00:54	
Control Test	0.081	00:54	
Air Blank	0.000	00:54	
Subject Sample	# 1	00:55	Subject Test Refused
Breath Volume	0.000	00:55	
Air Blank	0.000	00:55	
Control Test	0.081	00:55	
Air Blank	0.000	00:56	
Diagnostics Check	OK	00:56	

Data Download Date/Time: 06/29/2017 12:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,ALEX,
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: DERY,DENISE
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 05/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14968
DRY GAS EXPIRY DATE: 08/06/2018
VIOLATION CODE: DUI
ARREST TIME: 00:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:57

SUBJECT'S LAST NAME: MCCAFFREY

DATE/TIME OF TEST: 06/02/2017 01:28:10

Diagnostics Check		OK	01:21
Air Blank		0.000	01:22
Control Test		0.081	01:22
Air Blank		0.000	01:23
Subject Sample	# 1	0.205	01:23
Breath Volume		2.582	01:23
Air Blank		0.000	01:24
Air Blank		0.000	01:26
Subject Sample	# 2	0.196	01:26
Breath Volume		2.335	01:26
Air Blank		0.000	01:27
Control Test		0.080	01:27
Air Blank		0.000	01:28
Diagnostics Check		OK	01:28

Data Download Date/Time: 06/29/2017 12:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KEOHANE,KUNIKO,
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: KEOHANE,KUNIKO
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 05/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14968
DRY GAS EXPIRY DATE: 08/06/2018
VIOLATION CODE: DUI
ARREST TIME: 05:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:23

SUBJECT'S LAST NAME: SARZUELA

DATE/TIME OF TEST: 06/09/2017 06:13:06

Diagnostics Check	OK	06:06	
Air Blank	0.000	06:07	
Control Test	0.082	06:07	
Air Blank	0.000	06:08	
Subject Sample # 1	0.213	06:08	
Breath Volume	2.273	06:08	
Air Blank	0.000	06:09	
Air Blank	0.000	06:10	
Subject Sample # 2	0.217	06:11	
Breath Volume	2.007	06:11	
Air Blank	0.000	06:11	
Control Test	0.087	06:12	Interferent Detect
Air Blank		06:13	Purge Fail

Data Download Date/Time: 06/29/2017 12:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBISON,DAVID,E
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: SAPP,BRIAN
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 05/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14968
DRY GAS EXPIRY DATE: 08/06/2018
VIOLATION CODE: BUI
ARREST TIME: 20:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: DINDO

DATE/TIME OF TEST: 06/10/2017 21:52:12

Diagnostics Check	OK	21:49	
Air Blank	0.000	21:49	
Control Test	0.082	21:49	
Air Blank	0.000	21:50	
Subject Sample	# 1	21:50	Interferent Detect
Breath Volume	2.117	21:50	
Air Blank	0.000	21:51	
Control Test	0.076	21:51	Interferent Detect
Air Blank	0.000	21:52	

Data Download Date/Time: 06/29/2017 12:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBISON,DAVID,E
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: SAPP,BRIAN
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 05/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14968
DRY GAS EXPIRY DATE: 08/06/2018
VIOLATION CODE: BUI
ARREST TIME: 20:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: DINDO

DATE/TIME OF TEST: 06/10/2017 22:00:30

Diagnostics Check	OK	21:54
Air Blank	0.000	21:55
Control Test	0.082	21:55
Air Blank	0.000	21:55
Subject Sample # 1	0.158	21:56
Breath Volume	1.550	21:56
Air Blank	0.000	21:56
Air Blank	0.000	21:58
Subject Sample # 2	0.151	21:59
Breath Volume	1.234	21:59
Air Blank	0.000	21:59
Control Test	0.081	21:59
Air Blank	0.000	22:00
Diagnostics Check	OK	22:00

Data Download Date/Time: 06/29/2017 12:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,ALEX,
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: RODRIGUEZ,ALEX
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 05/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14968
DRY GAS EXPIRY DATE: 08/06/2018
VIOLATION CODE: DUI
ARREST TIME: 03:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:15

SUBJECT'S LAST NAME: RAND

DATE/TIME OF TEST: 06/11/2017 04:45:05

Diagnostics Check	OK	04:38
Air Blank	0.000	04:39
Control Test	0.082	04:39
Air Blank	0.000	04:40
Subject Sample # 1	0.198	04:40
Breath Volume	2.222	04:40
Air Blank	0.000	04:41
Air Blank	0.000	04:43
Subject Sample # 2	0.189	04:43
Breath Volume	1.371	04:43
Air Blank	0.000	04:44
Control Test	0.080	04:44
Air Blank	0.000	04:45
Diagnostics Check	OK	04:45

Data Download Date/Time: 06/29/2017 12:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELSMORE,JUSTIN,R
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: MOGERLEY,NEIL
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 05/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14968
DRY GAS EXPIRY DATE: 08/06/2018
VIOLATION CODE: DUI
ARREST TIME: 18:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:12

SUBJECT'S LAST NAME: SHADDUCK

DATE/TIME OF TEST: 06/11/2017 19:35:41

Diagnostics Check	OK	19:32	
Air Blank	0.000	19:33	
Control Test	0.082	19:33	
Air Blank	0.000	19:34	
Subject Sample	# 1	19:34	Subject Test Refused
Breath Volume	0.000	19:34	
Air Blank	0.000	19:34	
Control Test	0.082	19:35	
Air Blank	0.000	19:35	
Diagnostics Check	OK	19:35	

Data Download Date/Time: 06/29/2017 12:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,ALEX,
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: RODRIGUEZ,ALEX
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 05/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14968
DRY GAS EXPIRY DATE: 08/06/2018
VIOLATION CODE: DUI
ARREST TIME: 01:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:02

SUBJECT'S LAST NAME: BERRIOS

DATE/TIME OF TEST: 06/15/2017 02:30:20

Diagnostics Check	OK	02:27	
Air Blank	0.000	02:27	
Control Test	0.081	02:28	
Air Blank	0.000	02:28	
Subject Sample	# 1	02:28	Subject Test Refused
Breath Volume	0.000	02:28	
Air Blank	0.000	02:29	
Control Test	0.081	02:29	
Air Blank	0.000	02:30	
Diagnostics Check	OK	02:30	

Data Download Date/Time: 06/29/2017 12:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PERLMAN,GREGG,B
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: PERLMAN,GREGG
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 05/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14968
DRY GAS EXPIRY DATE: 08/06/2018
VIOLATION CODE: DUI
ARREST TIME: 14:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:00

SUBJECT'S LAST NAME: OJEDA

DATE/TIME OF TEST: 06/27/2017 17:36:46

Diagnostics Check	OK	17:31
Air Blank	0.000	17:31
Control Test	0.082	17:31
Air Blank	0.000	17:32
Subject Sample # 1	0.000	17:32
Breath Volume	1.828	17:32
Air Blank	0.000	17:33
Air Blank	0.000	17:35
Subject Sample # 2	0.000	17:35
Breath Volume	2.339	17:35
Air Blank	0.000	17:35
Control Test	0.080	17:36
Air Blank	0.000	17:36
Diagnostics Check	OK	17:36

Data Download Date/Time: 06/29/2017 12:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,ALEX,
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: CONATY,JAY
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 06/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14968
DRY GAS EXPIRY DATE: 08/06/2018
VIOLATION CODE: DUI
ARREST TIME: 23:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:32

SUBJECT'S LAST NAME: MOSHER

DATE/TIME OF TEST: 06/30/2017 00:00:58

Diagnostics Check	OK	23:57	
Air Blank	0.000	23:57	
Control Test	0.082	23:58	
Air Blank	0.000	23:58	
Subject Sample	# 1	23:59	Subject Test Refused
Breath Volume	0.000	23:59	
Air Blank	0.000	00:00	
Control Test	0.080	00:00	
Air Blank	0.000	00:00	
Diagnostics Check	OK	00:00	

Data Download Date/Time: 07/13/2017 17:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000872

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALOMINO,JULIO,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: HERNANDEZ,JASIEL
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 06/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 03:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:12

SUBJECT'S LAST NAME: RUFFINI

DATE/TIME OF TEST: 06/30/2017 04:53:39

Diagnostics Check	OK	04:50	
Air Blank	0.000	04:51	
Control Test	0.079	04:51	
Air Blank	0.000	04:51	
Subject Sample	# 1	04:52	Subject Test Refused
Breath Volume	0.000	04:52	
Air Blank	0.000	04:52	
Control Test	0.078	04:53	
Air Blank	0.000	04:53	
Diagnostics Check	OK	04:53	

Data Download Date/Time: 07/27/2017 14:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000873

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COSS,ATILA,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: COSS,ATILA
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 05/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 04:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: MEDINA

DATE/TIME OF TEST: 06/09/2017 04:54:02

Diagnostics Check	OK	04:51	
Air Blank	0.000	04:51	
Control Test	0.080	04:51	
Air Blank	0.000	04:52	
Subject Sample	# 1	04:52	Subject Test Refused
Breath Volume	0.000	04:52	
Air Blank	0.000	04:53	
Control Test	0.080	04:53	
Air Blank	0.000	04:53	
Diagnostics Check	OK	04:54	

Data Download Date/Time: 06/29/2017 15:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000873

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: JORVA,ALEXANDE,M

BREATH TEST OPERATOR AGENCY: Coral Gables PD

ARRESTING OFFICER: RAMBERT,MICHAEL

ARRESTING OFFICER AGENCY: Coral Gables PD

LAST AGENCY INSPECTION: 05/30/2017

Date or Time Changed: No

DRY GAS STANDARD LOT #: 203498

DRY GAS EXPIRY DATE: 12/10/2018

VIOLATION CODE: DUI

ARREST TIME: 06:30

GENDER: Female

OBSERVATION PERIOD BEGAN: 08:02

SUBJECT'S LAST NAME: WANSHEL

DATE/TIME OF TEST: 06/10/2017 08:25:12

Diagnostics Check	OK	08:22	
Air Blank	0.000	08:22	
Control Test	0.079	08:23	
Air Blank	0.000	08:23	
Subject Sample	# 1	08:23	Subject Test Refused
Breath Volume	0.000	08:23	
Air Blank	0.000	08:24	
Control Test	0.080	08:24	
Air Blank	0.000	08:25	
Diagnostics Check	OK	08:25	

Data Download Date/Time: 06/29/2017 15:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000873

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAMIREZ,MOISES,A
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: RAMIREZ,MOISES
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 05/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:36

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 06/11/2017 02:08:11

Diagnostics Check	OK	02:05	
Air Blank	0.000	02:05	
Control Test	0.080	02:06	
Air Blank	0.000	02:06	
Subject Sample	# 1	02:06	Subject Test Refused
Breath Volume	0.000	02:06	
Air Blank	0.000	02:07	
Control Test	0.081	02:07	
Air Blank	0.000	02:08	
Diagnostics Check	OK	02:08	

Data Download Date/Time: 06/29/2017 15:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000873

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUES,LUCAS,D
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: RODRIGUES,LUCAS
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 05/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 00:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:21

SUBJECT'S LAST NAME: MOREJON FUENTES

DATE/TIME OF TEST: 06/12/2017 00:49:42

Diagnostics Check	OK	00:46	
Air Blank	0.000	00:47	
Control Test	0.080	00:47	
Air Blank	0.000	00:47	
Subject Sample	# 1	00:48	Subject Test Refused
Breath Volume	0.000	00:48	
Air Blank	0.000	00:48	
Control Test	0.081	00:49	
Air Blank	0.000	00:49	
Diagnostics Check	OK	00:49	

Data Download Date/Time: 06/29/2017 15:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000873

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JORVA,ALEXANDER,M
BREATH TEST OPERATOR AGENCY: Coral Gables PD
ARRESTING OFFICER: HARZINSKI,CHRISTOPHER
ARRESTING OFFICER AGENCY: Coral Gables PD
LAST AGENCY INSPECTION: 05/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 23:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:58

SUBJECT'S LAST NAME: KANUMURI

DATE/TIME OF TEST: 06/13/2017 01:27:19

Diagnostics Check		OK	01:18
Air Blank		0.000	01:18
Control Test		0.080	01:19
Air Blank		0.000	01:19
Subject Sample	# 1	0.166	01:21
Breath Volume		1.335	01:21
Air Blank		0.000	01:22
Air Blank		0.000	01:24
Subject Sample	# 2	0.174	01:25
Breath Volume		2.089	01:25
Air Blank		0.000	01:26
Control Test		0.079	01:26
Air Blank		0.000	01:27
Diagnostics Check		OK	01:27

Data Download Date/Time: 06/29/2017 15:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000873

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: JORVA,ALEXANDER,M

BREATH TEST OPERATOR AGENCY: Coral Gables PD

ARRESTING OFFICER: MARTIN,NATALIE

ARRESTING OFFICER AGENCY: Coral Gables PD

LAST AGENCY INSPECTION: 05/30/2017

Date or Time Changed: No

DRY GAS STANDARD LOT #: 203498

DRY GAS EXPIRY DATE: 12/10/2018

VIOLATION CODE: DUI

ARREST TIME: 01:51

GENDER: Female

OBSERVATION PERIOD BEGAN: 02:06

SUBJECT'S LAST NAME: AGUILERA

DATE/TIME OF TEST: 06/29/2017 02:29:31

Diagnostics Check	OK	02:26	
Air Blank	0.000	02:27	
Control Test	0.080	02:27	
Air Blank	0.000	02:27	
Subject Sample	# 1	02:28	Subject Test Refused
Breath Volume	0.000	02:28	
Air Blank	0.000	02:28	
Control Test	0.080	02:28	
Air Blank	0.000	02:29	
Diagnostics Check	OK	02:29	

Data Download Date/Time: 06/29/2017 15:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000873

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHADWELL,WENDY,B
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: CHADWELL,WENDY
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 05/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 01:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:17

SUBJECT'S LAST NAME: ENCINA

DATE/TIME OF TEST: 06/29/2017 02:46:06

Diagnostics Check	OK	02:39
Air Blank	0.000	02:39
Control Test	0.080	02:39
Air Blank	0.000	02:40
Subject Sample # 1	0.079	02:41
Breath Volume	2.144	02:41
Air Blank	0.000	02:42
Air Blank	0.000	02:44
Subject Sample # 2	0.075	02:44
Breath Volume	2.460	02:44
Air Blank	0.000	02:45
Control Test	0.079	02:45
Air Blank	0.000	02:46
Diagnostics Check	OK	02:46

Data Download Date/Time: 06/29/2017 15:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000875

INSTRUMENT REGISTERED TO: CALHOUN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LINCOLN,TREY,B
BREATH TEST OPERATOR AGENCY: Calhoun County SO
ARRESTING OFFICER: LINCOLN,TREY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 05/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A3
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: GRISSET

DATE/TIME OF TEST: 06/04/2017 01:32:07

Diagnostics Check	OK	01:25
Air Blank	0.000	01:26
Control Test	0.078	01:26
Air Blank	0.000	01:26
Subject Sample # 1	0.000	01:27
Breath Volume	3.933	01:27
Air Blank	0.000	01:28
Air Blank	0.000	01:30
Subject Sample # 2	0.000	01:30
Breath Volume	4.316	01:30
Air Blank	0.000	01:31
Control Test	0.078	01:31
Air Blank	0.000	01:32
Diagnostics Check	OK	01:32

Data Download Date/Time: 06/06/2017 10:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000875

INSTRUMENT REGISTERED TO: CALHOUN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRIMES,DENISE,M
BREATH TEST OPERATOR AGENCY: Calhoun County SO
ARRESTING OFFICER: GRIMES,DENISE
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 05/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A3
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: SELLERS

DATE/TIME OF TEST: 06/04/2017 21:27:27

Diagnostics Check	OK	21:21
Air Blank	0.000	21:21
Control Test	0.079	21:22
Air Blank	0.000	21:22
Subject Sample # 1	0.000	21:23
Breath Volume	4.621	21:23
Air Blank	0.000	21:23
Air Blank	0.000	21:25
Subject Sample # 2	0.000	21:26
Breath Volume	3.851	21:26
Air Blank	0.000	21:26
Control Test	0.079	21:26
Air Blank	0.000	21:27
Diagnostics Check	OK	21:27

Data Download Date/Time: 06/06/2017 10:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000876

INSTRUMENT REGISTERED TO: HIALEAH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOPEZ,CRISTIAN,
BREATH TEST OPERATOR AGENCY: Medley PD
ARRESTING OFFICER: LAGOS,GUILLERMO
ARRESTING OFFICER AGENCY: Medley PD
LAST AGENCY INSPECTION: 05/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 712031
DRY GAS EXPIRY DATE: 02/21/2020
VIOLATION CODE: DUI
ARREST TIME: 21:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 06/16/2017 22:31:51

Diagnostics Check	OK	22:28	
Air Blank	0.000	22:28	
Control Test	0.079	22:29	
Air Blank	0.000	22:29	
Subject Sample	# 1	22:30	Subject Test Refused
Breath Volume	0.000	22:30	
Air Blank	0.000	22:30	
Control Test	0.079	22:31	
Air Blank	0.000	22:31	
Diagnostics Check	OK	22:31	

Data Download Date/Time: 06/22/2017 12:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAIGE,LEON,I
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: PAIGE,LEON
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 05/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1826916
DRY GAS EXPIRY DATE: 04/19/2019
VIOLATION CODE: DUI
ARREST TIME: 05:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:05

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 06/04/2017 06:37:03

Diagnostics Check	OK	06:34	
Air Blank	0.000	06:34	
Control Test	0.081	06:34	
Air Blank	0.000	06:35	
Subject Sample	# 1	06:35	Subject Test Refused
Breath Volume	0.000	06:35	
Air Blank	0.000	06:36	
Control Test	0.081	06:36	
Air Blank	0.000	06:36	
Diagnostics Check	OK	06:37	

Data Download Date/Time: 06/29/2017 13:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SOCARRAS,PEDRO,A
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: RIESGO,LAZARO
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 05/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1826916
DRY GAS EXPIRY DATE: 03/09/2018
VIOLATION CODE: DUI
ARREST TIME: 09:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:55

SUBJECT'S LAST NAME: SKALET

DATE/TIME OF TEST: 06/07/2017 10:22:49

Diagnostics Check	OK	10:15
Air Blank	0.000	10:16
Control Test	0.081	10:16
Air Blank	0.000	10:16
Subject Sample # 1	0.000	10:18
Breath Volume	1.398	10:18
Air Blank	0.000	10:18
Air Blank	0.000	10:20
Subject Sample # 2	0.000	10:21
Breath Volume	1.355	10:21
Air Blank	0.000	10:21
Control Test	0.081	10:22
Air Blank	0.000	10:22
Diagnostics Check	OK	10:22

Data Download Date/Time: 06/29/2017 13:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAIGE,LEON,I
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: PAIGE,LEON
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 05/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1826916
DRY GAS EXPIRY DATE: 04/19/2019
VIOLATION CODE: DUI
ARREST TIME: 02:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: PULCIANI

DATE/TIME OF TEST: 06/09/2017 03:03:54

Diagnostics Check	OK	02:57
Air Blank	0.000	02:57
Control Test	0.081	02:58
Air Blank	0.000	02:58
Subject Sample # 1	0.141	02:59
Breath Volume	2.562	02:59
Air Blank	0.000	03:00
Air Blank	0.000	03:02
Subject Sample # 2	0.141	03:02
Breath Volume	3.183	03:02
Air Blank	0.000	03:03
Control Test	0.081	03:03
Air Blank	0.000	03:03
Diagnostics Check	OK	03:03

Data Download Date/Time: 06/29/2017 13:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAIGE,LEON,I
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: PAIGE,LEON
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 05/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1826916
DRY GAS EXPIRY DATE: 04/19/2019
VIOLATION CODE: DUI
ARREST TIME: 04:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:00

SUBJECT'S LAST NAME: CAMERANESI

DATE/TIME OF TEST: 06/10/2017 05:28:41

Diagnostics Check	OK	05:22
Air Blank	0.000	05:22
Control Test	0.082	05:22
Air Blank	0.000	05:23
Subject Sample # 1	0.132	05:23
Breath Volume	2.886	05:23
Air Blank	0.000	05:24
Air Blank	0.000	05:26
Subject Sample # 2	0.123	05:27
Breath Volume	1.898	05:27
Air Blank	0.000	05:27
Control Test	0.082	05:28
Air Blank	0.000	05:28
Diagnostics Check	OK	05:28

Data Download Date/Time: 06/29/2017 13:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CREWS,CURTIS,L
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: PAIGE,LEON
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 05/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1826916
DRY GAS EXPIRY DATE: 03/09/2018
VIOLATION CODE: DUI
ARREST TIME: 03:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:57

SUBJECT'S LAST NAME: TRUDELL

DATE/TIME OF TEST: 06/11/2017 04:26:19

Diagnostics Check	OK	04:19
Air Blank	0.000	04:20
Control Test	0.082	04:20
Air Blank	0.000	04:21
Subject Sample # 1	0.167	04:21
Breath Volume	1.144	04:21
Air Blank	0.000	04:22
Air Blank	0.000	04:24
Subject Sample # 2	0.162	04:24
Breath Volume	1.648	04:24
Air Blank	0.000	04:25
Control Test	0.080	04:25
Air Blank	0.000	04:26
Diagnostics Check	OK	04:26

Data Download Date/Time: 06/29/2017 13:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAIGE,LEON,I
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: PAIGE,LEON
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 05/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1826916
DRY GAS EXPIRY DATE: 04/19/2019
VIOLATION CODE: DUI
ARREST TIME: 06:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:40

SUBJECT'S LAST NAME: SOLANO OLAZABAL

DATE/TIME OF TEST: 06/11/2017 07:03:31

Diagnostics Check	OK	07:00	
Air Blank	0.000	07:01	
Control Test	0.081	07:01	
Air Blank	0.000	07:01	
Subject Sample	# 1	07:02	RFI Detect
Breath Volume	0.000	07:02	
Air Blank	0.000	07:02	
Control Test	0.081	07:02	
Air Blank	0.000	07:03	
Diagnostics Check	OK	07:03	

Data Download Date/Time: 06/29/2017 13:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAIGE,LEON,I
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: PAIGE,LEON
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 05/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1826916
DRY GAS EXPIRY DATE: 03/09/2018
VIOLATION CODE: DUI
ARREST TIME: 06:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:40

SUBJECT'S LAST NAME: SOLANO OLAZABAL

DATE/TIME OF TEST: 06/11/2017 07:11:59

Diagnostics Check	OK	07:05
Air Blank	0.000	07:05
Control Test	0.081	07:06
Air Blank	0.000	07:06
Subject Sample # 1	0.214	07:07
Breath Volume	2.507	07:07
Air Blank	0.000	07:08
Air Blank	0.000	07:09
Subject Sample # 2	0.215	07:10
Breath Volume	3.648	07:10
Air Blank	0.000	07:11
Control Test	0.081	07:11
Air Blank	0.000	07:11
Diagnostics Check	OK	07:11

Data Download Date/Time: 06/29/2017 13:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,JOSE,D
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: PEREZ,JOSE
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 05/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1826916
DRY GAS EXPIRY DATE: 03/09/2018
VIOLATION CODE: DUI
ARREST TIME: 18:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:31

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 06/11/2017 19:05:40

Diagnostics Check	OK	18:59
Air Blank	0.000	18:59
Control Test	0.082	19:00
Air Blank	0.000	19:00
Subject Sample # 1	0.136	19:01
Breath Volume	1.792	19:01
Air Blank	0.000	19:01
Air Blank	0.000	19:03
Subject Sample # 2	0.138	19:04
Breath Volume	2.066	19:04
Air Blank	0.000	19:04
Control Test	0.081	19:05
Air Blank	0.000	19:05
Diagnostics Check	OK	19:05

Data Download Date/Time: 06/29/2017 13:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAIGE,LEON,I
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: PAIGE,LEON
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 05/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1826916
DRY GAS EXPIRY DATE: 04/19/2019
VIOLATION CODE: DUI
ARREST TIME: 04:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: FERRA

DATE/TIME OF TEST: 06/18/2017 04:36:32

Diagnostics Check	OK	04:33	
Air Blank	0.000	04:34	
Control Test	0.082	04:34	
Air Blank	0.000	04:34	
Subject Sample	# 1	04:35	Subject Test Refused
Breath Volume	0.000	04:35	
Air Blank	0.000	04:35	
Control Test	0.081	04:35	
Air Blank	0.000	04:36	
Diagnostics Check	OK	04:36	

Data Download Date/Time: 06/29/2017 13:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLAN,KEVIN,
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: PAIGE,LEON
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 05/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1826916
DRY GAS EXPIRY DATE: 03/09/2018
VIOLATION CODE: DUI
ARREST TIME: 06:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:48

SUBJECT'S LAST NAME: HERNANDEZ JR

DATE/TIME OF TEST: 06/19/2017 07:16:58

Diagnostics Check	OK	07:10
Air Blank	0.000	07:10
Control Test	0.082	07:11
Air Blank	0.000	07:11
Subject Sample # 1	0.000	07:12
Breath Volume	2.429	07:12
Air Blank	0.000	07:13
Air Blank	0.000	07:15
Subject Sample # 2	0.000	07:15
Breath Volume	2.246	07:15
Air Blank	0.000	07:16
Control Test	0.081	07:16
Air Blank	0.000	07:16
Diagnostics Check	OK	07:16

Data Download Date/Time: 06/29/2017 13:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAIGE,LEON,I
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: PAIGE,LEON
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 05/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1826916
DRY GAS EXPIRY DATE: 04/19/2019
VIOLATION CODE: DUI
ARREST TIME: 00:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: PACHECO

DATE/TIME OF TEST: 06/25/2017 01:14:25

Diagnostics Check	OK	01:11	
Air Blank	0.000	01:11	
Control Test	0.081	01:12	
Air Blank	0.000	01:12	
Subject Sample	# 1	01:13	Subject Test Refused
Breath Volume	0.000	01:13	
Air Blank	0.000	01:13	
Control Test	0.081	01:13	
Air Blank	0.000	01:14	
Diagnostics Check	OK	01:14	

Data Download Date/Time: 06/29/2017 13:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONGE,GRETTEL,M
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: MONGE,GRETTEL
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 06/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1826916
DRY GAS EXPIRY DATE: 03/09/2018
VIOLATION CODE: DUI
ARREST TIME: 23:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: ALVARADO

DATE/TIME OF TEST: 06/30/2017 00:27:32

Diagnostics Check	OK	00:25	
Air Blank	0.000	00:26	
Control Test	0.081	00:26	
Air Blank	0.000	00:26	
		00:27	RFI Detect
Air Blank	0.000	00:27	

Data Download Date/Time: 08/07/2017 10:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONGE,GRETTEL,M
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: MONGE,GRETTEL
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 06/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1826916
DRY GAS EXPIRY DATE: 03/09/2018
VIOLATION CODE: DUI
ARREST TIME: 23:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: ALVARADO

DATE/TIME OF TEST: 06/30/2017 00:33:19

Diagnostics Check	OK	00:30	
Air Blank	0.000	00:30	
Control Test	0.080	00:31	
Air Blank	0.000	00:31	
Subject Sample	# 1	00:31	Subject Test Refused
Breath Volume	0.000	00:31	
Air Blank	0.000	00:32	
Control Test	0.081	00:32	
Air Blank		00:32	RFI Detect
Air Blank	0.000	00:33	

Data Download Date/Time: 08/07/2017 10:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERNANDEZ,JAVIER,
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: CALLAN,TONY
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 06/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1826916
DRY GAS EXPIRY DATE: 03/09/2018
VIOLATION CODE: DUI
ARREST TIME: 23:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: ANTELO

DATE/TIME OF TEST: 06/30/2017 23:52:54

Diagnostics Check	OK	23:46
Air Blank	0.000	23:47
Control Test	0.081	23:47
Air Blank	0.000	23:48
Subject Sample # 1	0.000	23:48
Breath Volume	2.117	23:48
Air Blank	0.000	23:49
Air Blank	0.000	23:51
Subject Sample # 2	0.000	23:51
Breath Volume	3.000	23:51
Air Blank	0.000	23:52
Control Test	0.080	23:52
Air Blank	0.000	23:52
Diagnostics Check	OK	23:52

Data Download Date/Time: 08/07/2017 10:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JANVIER,FRITZ,G
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: JANVIER,FRITZ
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:49

SUBJECT'S LAST NAME: LESTER III

DATE/TIME OF TEST: 06/17/2017 03:16:48

Diagnostics Check	OK	03:13	
Air Blank	0.000	03:14	
Control Test	0.079	03:14	
Air Blank	0.000	03:15	
Subject Sample	# 1	03:15	Subject Test Refused
Breath Volume	0.000	03:15	
Air Blank	0.000	03:15	
Control Test	0.081	03:16	
Air Blank	0.000	03:16	
Diagnostics Check	OK	03:16	

Data Download Date/Time: 07/18/2017 12:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,VIDESH,S
BREATH TEST OPERATOR AGENCY: Aventura PD
ARRESTING OFFICER: BARRETT,L
ARRESTING OFFICER AGENCY: Aventura PD
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:20

SUBJECT'S LAST NAME: MOONEERAM

DATE/TIME OF TEST: 06/17/2017 05:43:45

Diagnostics Check	OK	05:40	
Air Blank	0.000	05:41	
Control Test	0.078	05:41	
Air Blank	0.000	05:42	
Subject Sample	# 1	05:42	Subject Test Refused
Breath Volume	0.000	05:42	
Air Blank	0.000	05:42	
Control Test	0.078	05:43	
Air Blank	0.000	05:43	
Diagnostics Check	OK	05:43	

Data Download Date/Time: 07/18/2017 12:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELGADO,M,R
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: DELGADO,M
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 07:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:30

SUBJECT'S LAST NAME: GABRIEL

DATE/TIME OF TEST: 06/18/2017 09:03:29

Diagnostics Check	OK	09:02	
Air Blank	0.000	09:02	
Control Test	0.064	09:02	Control Outside Tolerance
Air Blank	0.000	09:03	

Data Download Date/Time: 07/18/2017 12:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELGADO,MIGUEL,R
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: DELGADO,MIGUEL
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 07:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:20

SUBJECT'S LAST NAME: GABRIEL

DATE/TIME OF TEST: 06/18/2017 09:22:15

Diagnostics Check	OK	09:20	
Air Blank	0.000	09:21	
Control Test	0.180	09:21	Control Outside Tolerance
Air Blank	0.000	09:22	

Data Download Date/Time: 07/18/2017 12:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000881

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIVERO,JOSE,I
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: JOSEPH,NELLY
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:52

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 06/23/2017 23:33:54

Diagnostics Check	OK	23:27
Air Blank	0.000	23:27
Control Test	0.078	23:28
Air Blank	0.000	23:28
Subject Sample # 1	0.256	23:29
Breath Volume	3.367	23:29
Air Blank	0.000	23:30
Air Blank	0.000	23:31
Subject Sample # 2	0.242	23:32
Breath Volume	2.410	23:32
Air Blank	0.000	23:32
Control Test	0.076	23:33
Air Blank	0.000	23:33
Diagnostics Check	OK	23:33

Data Download Date/Time: 07/18/2017 12:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000881

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,STEVEN,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: ADAMS,STEVEN
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: CASTRILLO

DATE/TIME OF TEST: 06/28/2017 04:29:21

Diagnostics Check	OK	04:23
Air Blank	0.000	04:23
Control Test	0.080	04:23
Air Blank	0.000	04:24
Subject Sample # 1	0.281	04:24
Breath Volume	3.031	04:24
Air Blank	0.000	04:25
Air Blank	0.000	04:27
Subject Sample # 2	0.280	04:27
Breath Volume	2.996	04:27
Air Blank	0.000	04:28
Control Test	0.080	04:28
Air Blank	0.000	04:29
Diagnostics Check	OK	04:29

Data Download Date/Time: 07/18/2017 12:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MESA,JONATHAN,A
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: MESA,JONATHAN
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 05/03/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 06:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:10

SUBJECT'S LAST NAME: SINGER CERTAD

DATE/TIME OF TEST: 06/04/2017 07:39:24

Diagnostics Check	OK	07:30
Air Blank	0.000	07:31
Control Test	0.080	07:31
Air Blank	0.000	07:31
Subject Sample # 1	0.119	07:34
Breath Volume	1.316	07:34
Air Blank	0.000	07:35
Air Blank	0.000	07:36
Subject Sample # 2	0.124	07:37
Breath Volume	1.222	07:37
Air Blank	0.000	07:38
Control Test	0.079	07:38
Air Blank	0.000	07:39
Diagnostics Check	OK	07:39

Data Download Date/Time: 06/22/2017 14:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: CLOSIUS,RICHARD
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 05/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:24

SUBJECT'S LAST NAME: CRUZ-VAZQUEZ

DATE/TIME OF TEST: 06/02/2017 03:55:53

Diagnostics Check		OK	03:49
Air Blank		0.000	03:50
Control Test		0.080	03:50
Air Blank		0.000	03:51
Subject Sample	# 1	0.034	03:51
Breath Volume		2.300	03:51
Air Blank		0.000	03:52
Air Blank		0.000	03:53
Subject Sample	# 2	0.036	03:54
Breath Volume		2.144	03:54
Air Blank		0.000	03:54
Control Test		0.080	03:55
Air Blank		0.000	03:55
Diagnostics Check		OK	03:55

Data Download Date/Time: 06/29/2017 13:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARDELAS,MAYLEE,
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: BARDELAS,MAYLE
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 05/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: VIERADASILVACREMASCO

DATE/TIME OF TEST: 06/03/2017 03:03:20

Diagnostics Check	OK	02:57	
Air Blank	0.000	02:58	
Control Test	0.081	02:58	
Air Blank	0.000	02:58	
Subject Sample	# 1 0.194	03:02	Volume Not Met
Breath Volume	0.085	03:02	
Air Blank		03:02	Purge Fail
Air Blank	0.000	03:03	

Data Download Date/Time: 06/29/2017 13:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARDELAS,MAYLEE,
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: BARDELAS,MAYLEE
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 05/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: VIERA DA SILVA CREMA

DATE/TIME OF TEST: 06/03/2017 03:13:57

Diagnostics Check	OK	03:05
Air Blank	0.000	03:06
Control Test	0.076	03:06
Air Blank	0.000	03:07
Subject Sample # 1	0.198	03:08
Breath Volume	1.660	03:08
Air Blank	0.000	03:09
Air Blank	0.000	03:10
Subject Sample # 2	0.198	03:12
Breath Volume	1.746	03:12
Air Blank	0.000	03:13
Control Test	0.076	03:13
Air Blank	0.000	03:13
Diagnostics Check	OK	03:13

Data Download Date/Time: 06/29/2017 13:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURGOS,O,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: VARGAS,D
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 05/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:59

SUBJECT'S LAST NAME: CASADO MESA

DATE/TIME OF TEST: 06/04/2017 01:29:40

Diagnostics Check	OK	01:26	
Air Blank	0.000	01:27	
Control Test	0.081	01:27	
Air Blank	0.000	01:28	
Subject Sample	# 1	01:28	Subject Test Refused
Breath Volume	0.000	01:28	
Air Blank	0.000	01:28	
Control Test	0.080	01:29	
Air Blank	0.000	01:29	
Diagnostics Check	OK	01:29	

Data Download Date/Time: 06/29/2017 13:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 05/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:59

SUBJECT'S LAST NAME: CASTRO

DATE/TIME OF TEST: 06/04/2017 02:32:45

Diagnostics Check	OK	02:29	
Air Blank	0.000	02:30	
Control Test	0.081	02:30	
Air Blank	0.000	02:31	
Subject Sample	# 1	02:31	Subject Test Refused
Breath Volume	0.000	02:31	
Air Blank	0.000	02:31	
Control Test	0.080	02:32	
Air Blank	0.000	02:32	
Diagnostics Check	OK	02:32	

Data Download Date/Time: 06/29/2017 13:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURGOS,O,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: KINSEY-SMITH,F
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 05/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 05:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:34

SUBJECT'S LAST NAME: ROSADO

DATE/TIME OF TEST: 06/08/2017 06:00:33

Diagnostics Check	OK	05:55	
Air Blank	0.000	05:55	
Control Test	0.080	05:56	
Air Blank	0.000	05:56	
Subject Sample	# 1	05:59	Subject Test Refused
Breath Volume	0.000	05:59	
Air Blank	0.000	05:59	
Control Test	0.079	05:59	
Air Blank	0.000	06:00	
Diagnostics Check	OK	06:00	

Data Download Date/Time: 06/29/2017 13:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MESA,JONATHAN,A
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: MESA,JONATHAN
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 05/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: LARGAESPADA SALAZAR

DATE/TIME OF TEST: 06/10/2017 04:53:08

Diagnostics Check	OK	04:50	
Air Blank	0.000	04:50	
Control Test	0.080	04:51	
Air Blank	0.000	04:51	
Subject Sample	# 1	04:51	Subject Test Refused
Breath Volume	0.000	04:51	
Air Blank	0.000	04:52	
Control Test	0.080	04:52	
Air Blank	0.000	04:53	
Diagnostics Check	OK	04:53	

Data Download Date/Time: 06/29/2017 13:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINOJOSA,DANIEL,M
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: HINOJOSA,DANIEL
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 05/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 18:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:58

SUBJECT'S LAST NAME: JARVIS

DATE/TIME OF TEST: 06/10/2017 19:31:15

Diagnostics Check	OK	19:24
Air Blank	0.000	19:25
Control Test	0.079	19:25
Air Blank	0.000	19:26
Subject Sample # 1	0.110	19:26
Breath Volume	2.476	19:26
Air Blank	0.000	19:27
Air Blank	0.000	19:29
Subject Sample # 2	0.106	19:29
Breath Volume	3.031	19:29
Air Blank	0.000	19:30
Control Test	0.078	19:30
Air Blank	0.000	19:31
Diagnostics Check	OK	19:31

Data Download Date/Time: 06/29/2017 13:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: CLOSIUS,RICHARD
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 05/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:18

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 06/11/2017 00:47:52

Diagnostics Check	OK	00:42
Air Blank	0.000	00:42
Control Test	0.081	00:42
Air Blank	0.000	00:43
Subject Sample # 1	0.000	00:43
Breath Volume	1.757	00:43
Air Blank	0.000	00:44
Air Blank	0.000	00:46
Subject Sample # 2	0.000	00:46
Breath Volume	1.937	00:46
Air Blank	0.000	00:46
Control Test	0.079	00:47
Air Blank	0.000	00:47
Diagnostics Check	OK	00:47

Data Download Date/Time: 06/29/2017 13:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,EMILIE,M
BREATH TEST OPERATOR AGENCY: Sweetwater PD
ARRESTING OFFICER: LORENZO,YOSNIEL
ARRESTING OFFICER AGENCY: Sweetwater PD
LAST AGENCY INSPECTION: 05/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 17:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:23

SUBJECT'S LAST NAME: ELOSEGUI

DATE/TIME OF TEST: 06/17/2017 19:05:30

Diagnostics Check	OK	18:52	
Air Blank	0.000	18:53	
Control Test	0.081	18:53	
Air Blank	0.000	18:54	
Subject Sample	# 1 0.163	18:54	
Breath Volume	1.269	18:54	
Air Blank	0.000	18:55	
Air Blank	0.000	18:57	
Subject Sample	# 2 0.168	19:00	Volume Not Met
Breath Volume	0.980	19:00	
Air Blank	0.000	19:01	
Air Blank	0.000	19:03	
Subject Sample	# 3 0.165	19:03	
Breath Volume	1.386	19:03	
Air Blank	0.000	19:04	
Control Test	0.076	19:04	
Air Blank	0.000	19:05	
Diagnostics Check	OK	19:05	

Data Download Date/Time: 06/29/2017 13:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINOJOSA,DANIEL,M
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: HINOJOSA,DANIEL
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 05/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 19:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 06/17/2017 20:28:11

Diagnostics Check	OK	20:21
Air Blank	0.000	20:22
Control Test	0.079	20:22
Air Blank	0.000	20:23
Subject Sample # 1	0.158	20:23
Breath Volume	2.503	20:23
Air Blank	0.000	20:24
Air Blank	0.000	20:26
Subject Sample # 2	0.155	20:26
Breath Volume	2.660	20:26
Air Blank	0.000	20:27
Control Test	0.079	20:27
Air Blank	0.000	20:28
Diagnostics Check	OK	20:28

Data Download Date/Time: 06/29/2017 13:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARDELAS,MAYLEE,
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: BARDELAS,MAYLE
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 05/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: PIRES-CAMPOS

DATE/TIME OF TEST: 06/18/2017 03:02:14

Diagnostics Check		OK	02:55
Air Blank		0.000	02:55
Control Test		0.080	02:56
Air Blank		0.000	02:56
Subject Sample	# 1	0.217	02:57
Breath Volume		2.011	02:57
Air Blank		0.000	02:57
Air Blank		0.000	02:59
Subject Sample	# 2	0.217	03:00
Breath Volume		2.613	03:00
Air Blank		0.000	03:01
Control Test		0.079	03:01
Air Blank		0.000	03:02
Diagnostics Check		OK	03:02

Data Download Date/Time: 06/29/2017 13:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: FRANCO,GREGORY
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 05/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:44

SUBJECT'S LAST NAME: LOPEZ-JARAMILLO

DATE/TIME OF TEST: 06/18/2017 04:10:53

Diagnostics Check	OK	04:05
Air Blank	0.000	04:05
Control Test	0.080	04:05
Air Blank	0.000	04:06
Subject Sample # 1	0.000	04:06
Breath Volume	1.628	04:06
Air Blank	0.000	04:07
Air Blank	0.000	04:09
Subject Sample # 2	0.000	04:09
Breath Volume	2.207	04:09
Air Blank	0.000	04:09
Control Test	0.080	04:10
Air Blank	0.000	04:10
Diagnostics Check	OK	04:10

Data Download Date/Time: 06/29/2017 13:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PLANAS,JUAN,C
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: ACOSTA,ALBERTO
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 05/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 08:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:05

SUBJECT'S LAST NAME: DOMINGUEZ PADRON

DATE/TIME OF TEST: 06/18/2017 10:38:15

Diagnostics Check	OK	10:35	
Air Blank	0.000	10:35	
Control Test	0.079	10:35	
Air Blank	0.000	10:36	
Subject Sample	# 1	10:36	Subject Test Refused
Breath Volume	0.000	10:36	
Air Blank	0.000	10:37	
Control Test	0.079	10:37	
Air Blank	0.000	10:38	
Diagnostics Check	OK	10:38	

Data Download Date/Time: 06/29/2017 13:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MESA,JONATHAN,A
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: MESA,JONATHAN
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 05/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:28

SUBJECT'S LAST NAME: LASTRE LEON

DATE/TIME OF TEST: 06/19/2017 00:58:14

Diagnostics Check		OK	00:48
Air Blank		0.000	00:49
Control Test		0.079	00:49
Air Blank		0.000	00:50
Subject Sample	# 1	0.188	00:50
Breath Volume		2.191	00:50
Air Blank		0.000	00:51
Air Blank		0.000	00:53
Subject Sample	# 2	0.167	00:53
Breath Volume		1.359	00:53
Air Blank		0.000	00:54
Air Blank		0.000	00:56
Subject Sample	# 3	0.181	00:56
Breath Volume		1.250	00:56
Air Blank		0.000	00:57
Control Test		0.079	00:57
Air Blank		0.000	00:58
Diagnostics Check		OK	00:58

Data Download Date/Time: 06/29/2017 13:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,EMILIE,M
BREATH TEST OPERATOR AGENCY: Sweetwater PD
ARRESTING OFFICER: LORENZO,YOSNIEL
ARRESTING OFFICER AGENCY: Sweetwater PD
LAST AGENCY INSPECTION: 05/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 18:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:00

SUBJECT'S LAST NAME: CALMUNDO

DATE/TIME OF TEST: 06/23/2017 19:45:59

Diagnostics Check	OK	19:39
Air Blank	0.000	19:40
Control Test	0.080	19:40
Air Blank	0.000	19:40
Subject Sample # 1	0.218	19:41
Breath Volume	1.644	19:41
Air Blank	0.000	19:42
Air Blank	0.000	19:43
Subject Sample # 2	0.213	19:44
Breath Volume	1.945	19:44
Air Blank	0.000	19:45
Control Test	0.079	19:45
Air Blank	0.000	19:45
Diagnostics Check	OK	19:45

Data Download Date/Time: 06/29/2017 13:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINOJOSA,DANIEL,M
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: HINOJOSA,DANIEL
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 05/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 18:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:42

SUBJECT'S LAST NAME: URRUTIA

DATE/TIME OF TEST: 06/23/2017 20:12:02

Diagnostics Check	OK	20:05
Air Blank	0.000	20:06
Control Test	0.079	20:06
Air Blank	0.000	20:06
Subject Sample # 1	0.137	20:07
Breath Volume	2.503	20:07
Air Blank	0.000	20:08
Air Blank	0.000	20:09
Subject Sample # 2	0.132	20:10
Breath Volume	1.886	20:10
Air Blank	0.000	20:11
Control Test	0.078	20:11
Air Blank	0.000	20:11
Diagnostics Check	OK	20:12

Data Download Date/Time: 06/29/2017 13:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAPANES,R,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: TAPANES,R
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:28

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 06/16/2017 01:57:20

Diagnostics Check		OK	01:49
Air Blank		0.000	01:49
Control Test		0.079	01:49
Air Blank		0.000	01:50
Subject Sample	# 1	0.184	01:51
Breath Volume		2.726	01:51
Air Blank		0.000	01:52
Air Blank		0.000	01:54
Subject Sample	# 2	0.183	01:55
Breath Volume		2.992	01:55
Air Blank		0.000	01:56
Control Test		0.078	01:56
Air Blank		0.000	01:57
Diagnostics Check		OK	01:57

Data Download Date/Time: 07/28/2017 13:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VARGAS,D,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: TAPANES,R
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:08

SUBJECT'S LAST NAME: SIM

DATE/TIME OF TEST: 06/17/2017 01:33:09

Diagnostics Check	OK	01:30	
Air Blank	0.000	01:31	
Control Test	0.079	01:31	
Air Blank	0.000	01:32	
Subject Sample	# 1	01:32	RFI Detect
Breath Volume	0.000	01:32	
Air Blank		01:32	RFI Detect
Air Blank	0.000	01:33	

Data Download Date/Time: 07/28/2017 13:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VARGAS,D,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: TAPANES,R
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:08

SUBJECT'S LAST NAME: SIM

DATE/TIME OF TEST: 06/17/2017 01:42:32

Diagnostics Check	OK	01:35
Air Blank	0.000	01:36
Control Test	0.079	01:36
Air Blank	0.000	01:37
Subject Sample # 1	0.104	01:37
Breath Volume	1.601	01:37
Air Blank	0.000	01:38
Air Blank	0.000	01:40
Subject Sample # 2	0.102	01:40
Breath Volume	2.957	01:40
Air Blank	0.000	01:41
Control Test	0.079	01:41
Air Blank	0.000	01:42
Diagnostics Check	OK	01:42

Data Download Date/Time: 07/28/2017 13:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAPANES,R,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: TAPANES,R
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:19

SUBJECT'S LAST NAME: BLANCHARD

DATE/TIME OF TEST: 06/20/2017 02:43:16

Diagnostics Check	OK	02:39	
Air Blank	0.000	02:40	
Control Test	0.079	02:40	
Air Blank	0.000	02:41	
Subject Sample	# 1	02:41	RFI Detect
Breath Volume	0.000	02:41	
Air Blank	0.000	02:42	
Control Test	0.079	02:42	
Air Blank	0.000	02:43	
Diagnostics Check	OK	02:43	

Data Download Date/Time: 07/28/2017 13:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAPANES,R,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: TAPANES,R
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:19

SUBJECT'S LAST NAME: BLANCHARD

DATE/TIME OF TEST: 06/20/2017 02:52:19

Diagnostics Check	OK	02:45
Air Blank	0.000	02:45
Control Test	0.079	02:46
Air Blank	0.000	02:46
Subject Sample # 1	0.115	02:47
Breath Volume	1.941	02:47
Air Blank	0.000	02:48
Air Blank	0.000	02:50
Subject Sample # 2	0.116	02:50
Breath Volume	2.382	02:50
Air Blank	0.000	02:51
Control Test	0.079	02:51
Air Blank	0.000	02:52
Diagnostics Check	OK	02:52

Data Download Date/Time: 07/28/2017 13:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUEN,JACKIE,M
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: YUEN,JACKIE
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 05:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:00

SUBJECT'S LAST NAME: GUTIERREZ

DATE/TIME OF TEST: 06/20/2017 06:46:53

Diagnostics Check	OK	06:35	
Air Blank	0.000	06:35	
Control Test	0.078	06:36	
Air Blank	0.000	06:36	
Subject Sample	# 1	06:39	No Sample Provided
Breath Volume	0.000	06:39	
Air Blank	0.000	06:40	
Air Blank	0.000	06:42	
Subject Sample	# 2	06:45	No Sample Provided
Breath Volume	0.000	06:45	
Air Blank	0.000	06:45	
Control Test	0.079	06:46	
Air Blank	0.000	06:46	
Diagnostics Check	OK	06:46	

Data Download Date/Time: 07/28/2017 13:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUEN,JACKIE,M
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: YUEN,JACKIE
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 05:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:00

SUBJECT'S LAST NAME: GUTIERREZ

DATE/TIME OF TEST: 06/20/2017 06:54:18

Diagnostics Check	OK	06:48	
Air Blank	0.000	06:49	
Control Test	0.079	06:49	
Air Blank	0.000	06:50	
Subject Sample	# 1	06:53	No Sample Provided
Breath Volume	0.000	06:53	
Air Blank	0.000	06:53	
Air Blank	0.000	06:54	

Data Download Date/Time: 07/28/2017 13:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VARGAS,D,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: CHANG,B
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:22

SUBJECT'S LAST NAME: JIMENEZ - YORRO

DATE/TIME OF TEST: 06/25/2017 02:50:58

Diagnostics Check	OK	02:43
Air Blank	0.000	02:43
Control Test	0.079	02:44
Air Blank	0.000	02:44
Subject Sample # 1	0.122	02:45
Breath Volume	1.546	02:45
Air Blank	0.000	02:46
Air Blank	0.000	02:48
Subject Sample # 2	0.121	02:49
Breath Volume	1.667	02:49
Air Blank	0.000	02:50
Control Test	0.079	02:50
Air Blank	0.000	02:50
Diagnostics Check	OK	02:50

Data Download Date/Time: 07/28/2017 13:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SLIMAK,M,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: SLIMAK,M
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:56

SUBJECT'S LAST NAME: RIGGINS

DATE/TIME OF TEST: 06/30/2017 00:26:29

Diagnostics Check	OK	00:21	
Air Blank	0.000	00:21	
Control Test	0.079	00:22	
Air Blank	0.000	00:22	
Subject Sample	# 1	00:24	Subject Test Refused
Breath Volume	0.050	00:24	
Air Blank	0.000	00:25	
Control Test	0.080	00:25	
Air Blank	0.000	00:26	
Diagnostics Check	OK	00:26	

Data Download Date/Time: 07/28/2017 13:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MESA,JONATHAN,A
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: MESA,JONATHAN
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:26

SUBJECT'S LAST NAME: HOUP

DATE/TIME OF TEST: 06/21/2017 00:50:01

Diagnostics Check	OK	00:47	
Air Blank	0.000	00:47	
Control Test	0.078	00:47	
Air Blank	0.000	00:48	
Subject Sample	# 1	00:48	Subject Test Refused
Breath Volume	0.000	00:48	
Air Blank	0.000	00:49	
Control Test	0.078	00:49	
Air Blank	0.000	00:49	
Diagnostics Check	OK	00:50	

Data Download Date/Time: 07/28/2017 13:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,EDWIN,M
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: PEREZ,EDWIN
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 08:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:19

SUBJECT'S LAST NAME: FIGUEREDO

DATE/TIME OF TEST: 06/23/2017 09:46:52

Diagnostics Check	OK	09:40
Air Blank	0.000	09:40
Control Test	0.079	09:40
Air Blank	0.000	09:41
Subject Sample # 1	0.158	09:42
Breath Volume	1.851	09:42
Air Blank	0.000	09:43
Air Blank	0.000	09:44
Subject Sample # 2	0.167	09:45
Breath Volume	2.328	09:45
Air Blank	0.000	09:45
Control Test	0.077	09:46
Air Blank	0.000	09:46
Diagnostics Check	OK	09:46

Data Download Date/Time: 07/28/2017 13:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOO,BRANDON,R
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: YOUNG,BRYAN
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: BLACK

DATE/TIME OF TEST: 06/27/2017 00:19:36

Diagnostics Check	OK	00:16	
Air Blank	0.000	00:17	
Control Test	0.076	00:17	
Air Blank	0.000	00:18	
Subject Sample	# 1	00:18	Subject Test Refused
Breath Volume	0.000	00:18	
Air Blank	0.000	00:18	
Control Test	0.073	00:19	Control Outside Tolerance
Air Blank	0.000	00:19	

Data Download Date/Time: 07/28/2017 13:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZEVEDO,MIKE,C
BREATH TEST OPERATOR AGENCY: North Miami Beach PD
ARRESTING OFFICER: BESTARD,BRANDON
ARRESTING OFFICER AGENCY: North Miami Beach PD
LAST AGENCY INSPECTION: 05/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:12

SUBJECT'S LAST NAME: KOVYRSHYN

DATE/TIME OF TEST: 06/02/2017 01:40:56

Diagnostics Check	OK	01:32	
Air Blank	0.000	01:32	
Control Test	0.078	01:33	
Air Blank	0.000	01:33	
Subject Sample	# 1	01:36	No Sample Provided
Breath Volume	0.000	01:36	
Air Blank	0.000	01:37	
Air Blank	0.000	01:39	
Subject Sample	# 2	01:39	Subject Test Refused
Breath Volume	0.000	01:39	
Air Blank	0.000	01:40	
Control Test	0.079	01:40	
Air Blank	0.000	01:40	
Diagnostics Check	OK	01:40	

Data Download Date/Time: 06/28/2017 12:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRKUS,JOSHUA,B
BREATH TEST OPERATOR AGENCY: North Miami Beach PD
ARRESTING OFFICER: AZEVEDO,MICHAEL
ARRESTING OFFICER AGENCY: North Miami Beach PD
LAST AGENCY INSPECTION: 05/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:54

SUBJECT'S LAST NAME: KOVLYAEV

DATE/TIME OF TEST: 06/02/2017 03:26:23

Diagnostics Check		OK	03:14	
Air Blank		0.000	03:14	
Control Test		0.080	03:15	
Air Blank		0.000	03:15	
Subject Sample	# 1	0.158	03:18	Volume Not Met
Breath Volume		0.039	03:18	
Air Blank		0.000	03:19	
Air Blank		0.000	03:21	
Subject Sample	# 2	0.185	03:24	Volume Not Met
Breath Volume		0.093	03:24	
Air Blank		0.000	03:25	
Control Test		0.079	03:25	
Air Blank		0.000	03:26	
Diagnostics Check		OK	03:26	

Data Download Date/Time: 06/28/2017 12:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZEVEDO,MIKE,C
BREATH TEST OPERATOR AGENCY: North Miami Beach PD
ARRESTING OFFICER: FARINIK
ARRESTING OFFICER AGENCY: North Miami Beach PD
LAST AGENCY INSPECTION: 05/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:09

SUBJECT'S LAST NAME: HENRY

DATE/TIME OF TEST: 06/02/2017 05:36:51

Diagnostics Check	OK	05:33	
Air Blank	0.000	05:33	
Control Test	0.079	05:34	
Air Blank	0.000	05:34	
Subject Sample	# 1	05:35	Subject Test Refused
Breath Volume	0.000	05:35	
Air Blank	0.000	05:35	
Control Test	0.079	05:36	
Air Blank	0.000	05:36	
Diagnostics Check	OK	05:36	

Data Download Date/Time: 06/28/2017 12:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AJO,MICHAEL,D
BREATH TEST OPERATOR AGENCY: North Miami Beach PD
ARRESTING OFFICER: DARDEN,C
ARRESTING OFFICER AGENCY: North Miami Beach PD
LAST AGENCY INSPECTION: 05/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:48

SUBJECT'S LAST NAME: ARCHIBALD

DATE/TIME OF TEST: 06/07/2017 22:28:13

Diagnostics Check	OK	22:25	
Air Blank	0.000	22:25	
Control Test	0.079	22:26	
Air Blank	0.000	22:26	
Subject Sample	# 1	22:26	Subject Test Refused
Breath Volume	0.000	22:26	
Air Blank	0.000	22:27	
Control Test	0.080	22:27	
Air Blank	0.000	22:28	
Diagnostics Check	OK	22:28	

Data Download Date/Time: 06/28/2017 12:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MANN,CORA,L
BREATH TEST OPERATOR AGENCY: North Miami Beach PD
ARRESTING OFFICER: MANN,CORA
ARRESTING OFFICER AGENCY: North Miami Beach PD
LAST AGENCY INSPECTION: 05/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:04

SUBJECT'S LAST NAME: POLINSKI

DATE/TIME OF TEST: 06/08/2017 21:03:21

Diagnostics Check	OK	21:01	
Air Blank	0.000	21:01	
Control Test	0.080	21:02	
Air Blank	0.000	21:02	
Subject Sample	# 1	21:02	Sequence Aborted
Breath Volume	0.000	21:02	
Air Blank	0.000	21:03	

Data Download Date/Time: 06/28/2017 12:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MANN,CORA,L
BREATH TEST OPERATOR AGENCY: North Miami Beach PD
ARRESTING OFFICER: MANN,CORA
ARRESTING OFFICER AGENCY: North Miami Beach PD
LAST AGENCY INSPECTION: 05/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:04

SUBJECT'S LAST NAME: POLINSKI

DATE/TIME OF TEST: 06/08/2017 21:12:09

Diagnostics Check	OK	21:06	
Air Blank	0.000	21:06	
Control Test	0.079	21:07	
Air Blank	0.000	21:07	
Subject Sample	# 1	21:10	Subject Test Refused
Breath Volume	0.000	21:10	
Air Blank	0.000	21:11	
Control Test	0.080	21:11	
Air Blank	0.000	21:12	
Diagnostics Check	OK	21:12	

Data Download Date/Time: 06/28/2017 12:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRKUS,JOSHUA,B
BREATH TEST OPERATOR AGENCY: North Miami Beach PD
ARRESTING OFFICER: AZEVEDO,MICHAEL
ARRESTING OFFICER AGENCY: North Miami Beach PD
LAST AGENCY INSPECTION: 05/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:19

SUBJECT'S LAST NAME: ROJAS

DATE/TIME OF TEST: 06/09/2017 02:50:58

Diagnostics Check	OK	02:42	
Air Blank	0.000	02:42	
Control Test	0.081	02:42	
Air Blank	0.000	02:43	
Subject Sample	# 1	02:46	No Sample Provided
Breath Volume	0.000	02:46	
Air Blank	0.000	02:47	
Air Blank	0.000	02:49	
Subject Sample	# 2	02:49	Subject Test Refused
Breath Volume	0.000	02:49	
Air Blank	0.000	02:49	
Control Test	0.080	02:50	
Air Blank		02:50	RFI Detect
Air Blank		02:50	RFI Detect

Data Download Date/Time: 06/28/2017 12:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRKUS,JOSHUA,B
BREATH TEST OPERATOR AGENCY: North Miami Beach PD
ARRESTING OFFICER: AZEVEDO,MICHAEL
ARRESTING OFFICER AGENCY: North Miami Beach PD
LAST AGENCY INSPECTION: 05/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:11

SUBJECT'S LAST NAME: DAVILA

DATE/TIME OF TEST: 06/15/2017 02:50:12

Diagnostics Check		OK	02:36	
Air Blank		0.000	02:36	
Control Test		0.078	02:37	
Air Blank		0.000	02:37	
Subject Sample	# 1	0.238	02:40	Volume Not Met
Breath Volume		0.867	02:40	
Air Blank		0.000	02:41	
Air Blank		0.000	02:43	
Subject Sample	# 2	0.255	02:44	
Breath Volume		2.417	02:44	
Air Blank		0.000	02:45	
Air Blank		0.000	02:47	
Subject Sample	# 3	0.239	02:48	
Breath Volume		1.414	02:48	
Air Blank		0.000	02:49	
Control Test		0.077	02:49	
Air Blank		0.000	02:50	
Diagnostics Check		OK	02:50	

Data Download Date/Time: 06/28/2017 12:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRKUS,JOSHUA,B
BREATH TEST OPERATOR AGENCY: North Miami Beach PD
ARRESTING OFFICER: AZEVEDO,MICHAEL
ARRESTING OFFICER AGENCY: North Miami Beach PD
LAST AGENCY INSPECTION: 05/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:32

SUBJECT'S LAST NAME: BORRELL

DATE/TIME OF TEST: 06/16/2017 04:01:08

Diagnostics Check	OK	03:52
Air Blank	0.000	03:53
Control Test	0.080	03:53
Air Blank	0.000	03:54
Subject Sample # 1	0.080	03:56
Breath Volume	1.730	03:56
Air Blank	0.000	03:56
Air Blank	0.000	03:58
Subject Sample # 2	0.076	03:59
Breath Volume	1.363	03:59
Air Blank	0.000	04:00
Control Test	0.079	04:00
Air Blank	0.000	04:01
Diagnostics Check	OK	04:01

Data Download Date/Time: 06/28/2017 12:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRKUS,JOSHUA,B
BREATH TEST OPERATOR AGENCY: North Miami Beach PD
ARRESTING OFFICER: HAQ,F
ARRESTING OFFICER AGENCY: North Miami Beach PD
LAST AGENCY INSPECTION: 05/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:37

SUBJECT'S LAST NAME: DEMETRIADES

DATE/TIME OF TEST: 06/17/2017 01:04:12

Diagnostics Check	OK	00:57
Air Blank	0.000	00:58
Control Test	0.079	00:58
Air Blank	0.000	00:59
Subject Sample # 1	0.000	00:59
Breath Volume	1.507	00:59
Air Blank	0.000	01:00
Air Blank	0.000	01:02
Subject Sample # 2	0.000	01:02
Breath Volume	1.335	01:02
Air Blank	0.000	01:03
Control Test	0.078	01:03
Air Blank	0.000	01:04
Diagnostics Check	OK	01:04

Data Download Date/Time: 06/28/2017 12:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELGADO,MIGUEL,R
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: DEGADO,MIGUEL
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 05/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 07:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:20

SUBJECT'S LAST NAME: GABRIEL

DATE/TIME OF TEST: 06/18/2017 10:38:31

Diagnostics Check	OK	10:27
Air Blank	0.000	10:28
Control Test	0.080	10:28
Air Blank	0.000	10:28
Subject Sample # 1	0.282	10:29
Breath Volume	1.386	10:29
Air Blank	0.000	10:30
Air Blank	0.000	10:31
Subject Sample # 2	0.242	10:33
Breath Volume	1.304	10:33
Air Blank	0.000	10:34
Air Blank	0.000	10:36
Subject Sample # 3	0.270	10:36
Breath Volume	2.582	10:36
Air Blank	0.000	10:37
Control Test	0.077	10:37
Air Blank	0.000	10:38
Diagnostics Check	OK	10:38

Data Download Date/Time: 06/28/2017 12:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TANIS,MIKE,
BREATH TEST OPERATOR AGENCY: North Miami Beach PD
ARRESTING OFFICER: MENDEZ,MICHAEL
ARRESTING OFFICER AGENCY: North Miami Beach PD
LAST AGENCY INSPECTION: 05/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: CASTRO

DATE/TIME OF TEST: 06/21/2017 02:40:06

Diagnostics Check	OK	02:37	
Air Blank	0.000	02:37	
Control Test	0.079	02:38	
Air Blank	0.000	02:38	
Subject Sample	# 1	02:38	Subject Test Refused
Breath Volume	0.000	02:38	
Air Blank	0.000	02:39	
Control Test	0.080	02:39	
Air Blank	0.000	02:40	
Diagnostics Check	OK	02:40	

Data Download Date/Time: 06/28/2017 12:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELGADO,MIGUEL,R
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: DELGADO,MIGUEL
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 05/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 09:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:10

SUBJECT'S LAST NAME: SALAZAR JR

DATE/TIME OF TEST: 06/28/2017 10:41:11

Diagnostics Check	OK	10:35
Air Blank	0.000	10:35
Control Test	0.079	10:35
Air Blank	0.000	10:36
Subject Sample # 1	0.000	10:36
Breath Volume	3.289	10:36
Air Blank	0.000	10:37
Air Blank	0.000	10:39
Subject Sample # 2	0.000	10:39
Breath Volume	4.503	10:39
Air Blank	0.000	10:40
Control Test	0.078	10:40
Air Blank	0.000	10:41
Diagnostics Check	OK	10:41

Data Download Date/Time: 06/28/2017 12:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZEVEDO,MIKE,C
BREATH TEST OPERATOR AGENCY: North Miami Beach PD
ARRESTING OFFICER: AZEVEDO,MIKE
ARRESTING OFFICER AGENCY: North Miami Beach PD
LAST AGENCY INSPECTION: 06/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:49

SUBJECT'S LAST NAME: DANIEL

DATE/TIME OF TEST: 06/29/2017 00:16:27

Diagnostics Check	OK	00:09
Air Blank	0.000	00:10
Control Test	0.079	00:10
Air Blank	0.000	00:11
Subject Sample # 1	0.104	00:11
Breath Volume	1.578	00:11
Air Blank	0.000	00:12
Air Blank	0.000	00:13
Subject Sample # 2	0.101	00:14
Breath Volume	1.304	00:14
Air Blank	0.000	00:15
Control Test	0.079	00:15
Air Blank	0.000	00:16
Diagnostics Check	OK	00:16

Data Download Date/Time: 07/27/2017 21:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRKUS,JOSHUA,B
BREATH TEST OPERATOR AGENCY: North Miami Beach PD
ARRESTING OFFICER: AZEVEDO,MICHAEL
ARRESTING OFFICER AGENCY: North Miami Beach PD
LAST AGENCY INSPECTION: 06/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:39

SUBJECT'S LAST NAME: GORY

DATE/TIME OF TEST: 06/30/2017 01:08:19

Diagnostics Check	OK	00:59	
Air Blank	0.000	01:00	
Control Test	0.079	01:00	
Air Blank	0.000	01:01	
Subject Sample	# 1	01:04	No Sample Provided
Breath Volume	0.000	01:04	
Air Blank	0.000	01:04	
Air Blank	0.000	01:06	
Subject Sample	# 2	01:06	Subject Test Refused
Breath Volume	0.000	01:06	
Air Blank	0.000	01:07	
Control Test	0.080	01:07	
Air Blank	0.000	01:08	
Diagnostics Check	OK	01:08	

Data Download Date/Time: 07/27/2017 21:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRKUS,JOSHUA,B
BREATH TEST OPERATOR AGENCY: North Miami Beach PD
ARRESTING OFFICER: AZEVEDO,MICHAEL
ARRESTING OFFICER AGENCY: North Miami Beach PD
LAST AGENCY INSPECTION: 06/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:01

SUBJECT'S LAST NAME: ROMERO

DATE/TIME OF TEST: 06/30/2017 02:27:37

Diagnostics Check	OK	02:21
Air Blank	0.000	02:21
Control Test	0.079	02:21
Air Blank	0.000	02:22
Subject Sample # 1	0.072	02:23
Breath Volume	1.449	02:23
Air Blank	0.000	02:23
Air Blank	0.000	02:25
Subject Sample # 2	0.072	02:26
Breath Volume	1.332	02:26
Air Blank	0.000	02:26
Control Test	0.079	02:27
Air Blank	0.000	02:27
Diagnostics Check	OK	02:27

Data Download Date/Time: 07/27/2017 21:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FWC	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 507129
BREATH TEST OPERATOR: PIFER,JUSTIN,S	DRY GAS EXPIRY DATE: 08/04/2019
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)	VIOLATION CODE: BUI
ARRESTING OFFICER: MOBLEY,BRITTANY	ARREST TIME: 17:38
ARRESTING OFFICER AGENCY: FFWCC (Panama City)	GENDER: Male
LAST AGENCY INSPECTION: 05/11/2017	OBSERVATION PERIOD BEGAN: 17:38
SUBJECT'S LAST NAME: MILLER	DATE/TIME OF TEST: 06/02/2017 18:35:45

Diagnostics Check	OK	18:29
Air Blank	0.000	18:29
Control Test	0.080	18:29
Air Blank	0.000	18:30
Subject Sample # 1	0.063	18:31
Breath Volume	1.441	18:31
Air Blank	0.000	18:31
Air Blank	0.000	18:33
Subject Sample # 2	0.061	18:34
Breath Volume	1.890	18:34
Air Blank	0.000	18:34
Control Test	0.079	18:35
Air Blank	0.000	18:35
Diagnostics Check	OK	18:35

Data Download Date/Time: 06/12/2017 16:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FWC	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 507129
BREATH TEST OPERATOR: PIFER,JUSTIN,S	DRY GAS EXPIRY DATE: 08/04/2019
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)	VIOLATION CODE: BUI
ARRESTING OFFICER: PIFER,JUSTIN	ARREST TIME: 21:15
ARRESTING OFFICER AGENCY: FFWCC (Panama City)	GENDER: Male
LAST AGENCY INSPECTION: 05/11/2017	OBSERVATION PERIOD BEGAN: 21:15

SUBJECT'S LAST NAME: MUNDHENK DATE/TIME OF TEST: 06/02/2017 22:07:24

Diagnostics Check	OK	22:00
Air Blank	0.000	22:01
Control Test	0.079	22:01
Air Blank	0.000	22:02
Subject Sample # 1	0.133	22:02
Breath Volume	2.882	22:02
Air Blank	0.000	22:03
Air Blank	0.000	22:05
Subject Sample # 2	0.127	22:05
Breath Volume	2.121	22:05
Air Blank	0.000	22:06
Control Test	0.080	22:06
Air Blank	0.000	22:07
Diagnostics Check	OK	22:07

Data Download Date/Time: 06/12/2017 16:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FWC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOLNAR,JARROD,T
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: CORBIN,BRITTON
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 05/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: BUI
ARREST TIME: 14:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:37

SUBJECT'S LAST NAME: THORNTON

DATE/TIME OF TEST: 06/03/2017 15:06:02

Diagnostics Check	OK	15:03	
Air Blank	0.000	15:03	
Control Test	0.080	15:04	
Air Blank	0.000	15:04	
Subject Sample	# 1	15:04	Subject Test Refused
Breath Volume	0.000	15:04	
Air Blank	0.000	15:05	
Control Test	0.080	15:05	
Air Blank	0.000	15:05	
Diagnostics Check	OK	15:06	

Data Download Date/Time: 06/12/2017 16:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FWC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOLNAR,JARROD,A
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: ROBERSON,LOUIE
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 05/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: BUI
ARREST TIME: 15:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:01

SUBJECT'S LAST NAME: TECUMSEH

DATE/TIME OF TEST: 06/03/2017 15:33:37

Diagnostics Check	OK	15:27
Air Blank	0.000	15:28
Control Test	0.080	15:28
Air Blank	0.000	15:28
Subject Sample # 1	0.082	15:29
Breath Volume	1.875	15:29
Air Blank	0.000	15:29
Air Blank	0.000	15:31
Subject Sample # 2	0.070	15:32
Breath Volume	1.191	15:32
Air Blank	0.000	15:32
Control Test	0.080	15:33
Air Blank	0.000	15:33
Diagnostics Check	OK	15:33

Data Download Date/Time: 06/12/2017 16:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FWC

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: WHITE,REID,B

BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)

ARRESTING OFFICER: LONG,SEAN

ARRESTING OFFICER AGENCY: FFWCC (West Palm)

LAST AGENCY INSPECTION: 05/11/2017

Date or Time Changed: No

DRY GAS STANDARD LOT #: 507129

DRY GAS EXPIRY DATE: 08/04/2019

VIOLATION CODE: BUI

ARREST TIME: 16:00

GENDER: Male

OBSERVATION PERIOD BEGAN: 15:59

SUBJECT'S LAST NAME: STONE

DATE/TIME OF TEST: 06/03/2017 16:26:13

Diagnostics Check	OK	16:23	
Air Blank	0.000	16:23	
Control Test	0.081	16:24	
Air Blank	0.000	16:24	
Subject Sample	# 1	16:24	Subject Test Refused
Breath Volume	0.000	16:24	
Air Blank	0.000	16:25	
Control Test	0.081	16:25	
Air Blank	0.000	16:26	
Diagnostics Check	OK	16:26	

Data Download Date/Time: 06/12/2017 16:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FWC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITE,REID,B
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: KLEIS,KEVIN
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 05/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: BUI
ARREST TIME: 16:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:36

SUBJECT'S LAST NAME: PRIDMORE

DATE/TIME OF TEST: 06/03/2017 17:05:00

Diagnostics Check	OK	16:56
Air Blank	0.000	16:57
Control Test	0.080	16:57
Air Blank	0.000	16:58
Subject Sample # 1	0.145	16:59
Breath Volume	1.382	16:59
Air Blank	0.000	17:00
Air Blank	0.000	17:02
Subject Sample # 2	0.156	17:03
Breath Volume	1.382	17:03
Air Blank	0.000	17:04
Control Test	0.079	17:04
Air Blank	0.000	17:04
Diagnostics Check	OK	17:05

Data Download Date/Time: 06/12/2017 16:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FWC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOLNAR,JARROD,T
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: MOLNAR,JARROD
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 05/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: BUI
ARREST TIME: 16:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:43

SUBJECT'S LAST NAME: PAONE

DATE/TIME OF TEST: 06/03/2017 17:52:16

Diagnostics Check	OK	17:49	
Air Blank	0.000	17:49	
Control Test	0.081	17:50	
Air Blank	0.000	17:50	
Subject Sample	# 1	17:50	Subject Test Refused
Breath Volume	0.000	17:50	
Air Blank	0.000	17:51	
Control Test	0.080	17:51	
Air Blank	0.000	17:52	
Diagnostics Check	OK	17:52	

Data Download Date/Time: 06/12/2017 16:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FWC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORBIN,BRITTON,L
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: CORBIN,BRITTON
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 05/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: BUI
ARREST TIME: 17:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:45

SUBJECT'S LAST NAME: HOLZ

DATE/TIME OF TEST: 06/03/2017 18:28:24

Diagnostics Check	OK	18:25	
Air Blank	0.000	18:26	
Control Test	0.081	18:26	
Air Blank	0.000	18:26	
Subject Sample	# 1	18:27	Subject Test Refused
Breath Volume	0.000	18:27	
Air Blank	0.000	18:27	
Control Test	0.081	18:27	
Air Blank	0.000	18:28	
Diagnostics Check	OK	18:28	

Data Download Date/Time: 06/12/2017 16:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FWC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOLNAR,JARROD,T
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: ROBERSON,LOUIE
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 05/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: BUI
ARREST TIME: 18:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:15

SUBJECT'S LAST NAME: MILLS

DATE/TIME OF TEST: 06/03/2017 19:45:09

Diagnostics Check	OK	19:42	
Air Blank	0.000	19:42	
Control Test	0.081	19:43	
Air Blank	0.000	19:43	
Subject Sample	# 1	19:43	Subject Test Refused
Breath Volume	0.000	19:43	
Air Blank	0.000	19:44	
Control Test	0.081	19:44	
Air Blank	0.000	19:45	
Diagnostics Check	OK	19:45	

Data Download Date/Time: 06/12/2017 16:23

**Florida Department of Law Enforcement
Alcohol Testing Program**

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FWC	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 507129
BREATH TEST OPERATOR: PIFER,JUSTIN,S	DRY GAS EXPIRY DATE: 08/04/2019
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)	VIOLATION CODE: BUI
ARRESTING OFFICER: CORBIN,BRITTON	ARREST TIME: 15:22
ARRESTING OFFICER AGENCY: FFWCC (Panama City)	GENDER: Male
LAST AGENCY INSPECTION: 05/11/2017	OBSERVATION PERIOD BEGAN: 15:22

SUBJECT'S LAST NAME: MARCURE DATE/TIME OF TEST: 06/08/2017 16:11:39

Diagnostics Check	OK	16:08	
Air Blank	0.000	16:09	
Control Test	0.080	16:09	
Air Blank	0.000	16:09	
Subject Sample	# 1	16:10	Subject Test Refused
Breath Volume	0.000	16:10	
Air Blank	0.000	16:10	
Control Test	0.081	16:11	
Air Blank	0.000	16:11	
Diagnostics Check	OK	16:11	

Data Download Date/Time: 06/12/2017 16:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FWC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIFER,JUSTIN,S
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: BARTLETT,SHELTON
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 06/12/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: BUI
ARREST TIME: 17:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:46

SUBJECT'S LAST NAME: BURROWS 3

DATE/TIME OF TEST: 06/17/2017 18:36:00

Diagnostics Check	OK	18:29
Air Blank	0.000	18:29
Control Test	0.080	18:29
Air Blank	0.000	18:30
Subject Sample # 1	0.151	18:31
Breath Volume	3.046	18:31
Air Blank	0.000	18:32
Air Blank	0.000	18:33
Subject Sample # 2	0.142	18:34
Breath Volume	2.386	18:34
Air Blank	0.000	18:35
Control Test	0.076	18:35
Air Blank	0.000	18:35
Diagnostics Check	OK	18:36

Data Download Date/Time: 07/18/2017 14:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000904

INSTRUMENT REGISTERED TO: FFWCC SOUTH REGION
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MATTSON,CHRIS,J
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: GARCIA,NOEL
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 05/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1757137
DRY GAS EXPIRY DATE: 10/15/2017
VIOLATION CODE: BUI
ARREST TIME: 19:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: MILETTI

DATE/TIME OF TEST: 06/17/2017 20:27:06

Diagnostics Check	OK	20:21
Air Blank	0.000	20:21
Control Test	0.079	20:21
Air Blank	0.000	20:22
Subject Sample # 1	0.203	20:22
Breath Volume	3.023	20:22
Air Blank	0.000	20:23
Air Blank	0.000	20:25
Subject Sample # 2	0.195	20:25
Breath Volume	2.449	20:25
Air Blank	0.000	20:26
Control Test	0.079	20:26
Air Blank	0.000	20:27
Diagnostics Check	OK	20:27

Data Download Date/Time: 06/28/2017 11:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000907

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROWE,BRANDON,D
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: ROWE,BRANDON
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 05/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 02:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: MARTIN

DATE/TIME OF TEST: 06/10/2017 03:30:00

Diagnostics Check		OK	03:17	
Air Blank		0.000	03:17	
Control Test		0.078	03:17	
Air Blank		0.000	03:18	
Subject Sample	# 1	0.000	03:21	Volume Not Met
Breath Volume		0.160	03:21	
Air Blank		0.000	03:22	
Air Blank		0.000	03:23	
Subject Sample	# 2	0.220	03:24	
Breath Volume		1.988	03:24	
Air Blank		0.000	03:25	
Air Blank		0.000	03:27	
Subject Sample	# 3	0.211	03:28	
Breath Volume		1.851	03:28	
Air Blank		0.000	03:29	
Control Test		0.080	03:29	
Air Blank		0.000	03:29	
Diagnostics Check		OK	03:30	

Data Download Date/Time: 07/24/2017 09:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000907

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROWE,BRANDON,D
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: BRADLEY,ZACHARY
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 05/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 03:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:11

SUBJECT'S LAST NAME: NUNEZ VELA

DATE/TIME OF TEST: 06/11/2017 03:56:31

Diagnostics Check	OK	03:49
Air Blank	0.000	03:49
Control Test	0.079	03:50
Air Blank	0.000	03:50
Subject Sample # 1	0.130	03:52
Breath Volume	2.871	03:52
Air Blank	0.000	03:52
Air Blank	0.000	03:54
Subject Sample # 2	0.130	03:55
Breath Volume	1.675	03:55
Air Blank	0.000	03:55
Control Test	0.079	03:55
Air Blank	0.000	03:56
Diagnostics Check	OK	03:56

Data Download Date/Time: 07/24/2017 09:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000940

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRATSCH,CASEY,R
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: BRATSCH,CASEY
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 05/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:37

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 06/10/2017 04:20:29

Diagnostics Check	OK	04:17	
Air Blank	0.000	04:17	
Control Test	0.077	04:17	
Air Blank	0.000	04:18	
Subject Sample	# 1	04:19	Subject Test Refused
Breath Volume	0.000	04:19	
Air Blank	0.000	04:19	
Control Test	0.078	04:19	
Air Blank	0.000	04:20	
Diagnostics Check	OK	04:20	

Data Download Date/Time: 06/21/2017 19:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000940

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRATSCH,CASEY,R
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: BRATSCH,CASEY
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 05/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:42

SUBJECT'S LAST NAME: MINGCHAI

DATE/TIME OF TEST: 06/11/2017 03:09:00

Diagnostics Check	OK	03:02
Air Blank	0.000	03:03
Control Test	0.078	03:03
Air Blank	0.000	03:04
Subject Sample # 1	0.168	03:04
Breath Volume	2.093	03:04
Air Blank	0.000	03:05
Air Blank	0.000	03:06
Subject Sample # 2	0.162	03:07
Breath Volume	2.996	03:07
Air Blank	0.000	03:08
Control Test	0.079	03:08
Air Blank	0.000	03:08
Diagnostics Check	OK	03:09

Data Download Date/Time: 06/21/2017 19:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000940

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLEVELAND, WILLIAM, S
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: DRULARD, BRIAN
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 05/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:14

SUBJECT'S LAST NAME: MCGINNIS

DATE/TIME OF TEST: 06/13/2017 04:41:04

Diagnostics Check		OK	04:34
Air Blank		0.000	04:35
Control Test		0.079	04:35
Air Blank		0.000	04:35
Subject Sample	# 1	0.254	04:36
Breath Volume		2.906	04:36
Air Blank		0.000	04:37
Air Blank		0.000	04:38
Subject Sample	# 2	0.254	04:39
Breath Volume		2.769	04:39
Air Blank		0.000	04:40
Control Test		0.079	04:40
Air Blank		0.000	04:40
Diagnostics Check		OK	04:41

Data Download Date/Time: 06/21/2017 19:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000940

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRATSCH,CASEY,R
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: BRATSCH,CASEY
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 05/17/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: SHIPPY

DATE/TIME OF TEST: 06/15/2017 02:27:33

Diagnostics Check	OK	02:20
Air Blank	0.000	02:21
Control Test	0.078	02:21
Air Blank	0.000	02:22
Subject Sample # 1	0.137	02:22
Breath Volume	1.664	02:22
Air Blank	0.000	02:23
Air Blank	0.000	02:25
Subject Sample # 2	0.142	02:26
Breath Volume	1.285	02:26
Air Blank	0.000	02:26
Control Test	0.079	02:26
Air Blank	0.000	02:27
Diagnostics Check	OK	02:27

Data Download Date/Time: 06/21/2017 19:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000940

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,NICHOLAS,W
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: ELLER,GRANT
ARRESTING OFFICER AGENCY: FFWCC (Ocala)
LAST AGENCY INSPECTION: 05/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: BUI
ARREST TIME: 15:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:30

SUBJECT'S LAST NAME: YANKEE

DATE/TIME OF TEST: 06/17/2017 17:42:02

Diagnostics Check	OK	17:39	
Air Blank	0.000	17:39	
Control Test	0.078	17:40	
Air Blank	0.000	17:40	
Subject Sample	# 1	17:40	Subject Test Refused
Breath Volume	0.000	17:40	
Air Blank	0.000	17:41	
Control Test	0.078	17:41	
Air Blank	0.000	17:41	
Diagnostics Check	OK	17:42	

Data Download Date/Time: 06/21/2017 19:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000940

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWARD, WILLIAM,
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: BREWER, KYLE
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 05/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: VERGARA

DATE/TIME OF TEST: 06/19/2017 02:51:35

Diagnostics Check	OK	02:43	
Air Blank	0.000	02:43	
Control Test	0.080	02:44	
Air Blank	0.000	02:44	
Subject Sample # 1	0.185	02:46	
Breath Volume	1.574	02:46	
Air Blank	0.000	02:46	
Air Blank	0.000	02:48	
Subject Sample # 2		02:50	Subject Test Refused
Breath Volume	0.199	02:50	
Air Blank	0.000	02:50	
Control Test	0.077	02:51	
Air Blank	0.000	02:51	
Diagnostics Check	OK	02:51	

Data Download Date/Time: 06/21/2017 19:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MERTILUS,FREDELYNE,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: BRAWNER,M
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: HUFF

DATE/TIME OF TEST: 06/03/2017 00:31:58

Diagnostics Check	OK	00:25
Air Blank	0.000	00:26
Control Test	0.079	00:26
Air Blank	0.000	00:26
Subject Sample # 1	0.194	00:27
Breath Volume	2.394	00:27
Air Blank	0.000	00:28
Air Blank	0.000	00:30
Subject Sample # 2	0.192	00:30
Breath Volume	2.609	00:30
Air Blank	0.000	00:31
Control Test	0.078	00:31
Air Blank	0.000	00:31
Diagnostics Check	OK	00:31

Data Download Date/Time: 07/05/2017 15:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,MARTIN,C
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: MILLER,K
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:32

SUBJECT'S LAST NAME: PACHECO TORRES

DATE/TIME OF TEST: 06/03/2017 22:03:27

Diagnostics Check	OK	21:56
Air Blank	0.000	21:56
Control Test	0.079	21:56
Air Blank	0.000	21:57
Subject Sample # 1	0.063	21:58
Breath Volume	1.292	21:58
Air Blank	0.000	21:59
Air Blank	0.000	22:01
Subject Sample # 2	0.068	22:01
Breath Volume	1.292	22:01
Air Blank	0.000	22:02
Control Test	0.079	22:02
Air Blank	0.000	22:03
Diagnostics Check	OK	22:03

Data Download Date/Time: 07/05/2017 15:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANNORAT,ADELY,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: LEONI,S
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:24

SUBJECT'S LAST NAME: ROMAN

DATE/TIME OF TEST: 06/05/2017 03:00:54

Diagnostics Check	OK	02:54
Air Blank	0.000	02:55
Control Test	0.080	02:55
Air Blank	0.000	02:56
Subject Sample # 1	0.176	02:56
Breath Volume	2.070	02:56
Air Blank	0.000	02:57
Air Blank	0.000	02:58
Subject Sample # 2	0.180	02:59
Breath Volume	2.628	02:59
Air Blank	0.000	03:00
Control Test	0.080	03:00
Air Blank	0.000	03:00
Diagnostics Check	OK	03:00

Data Download Date/Time: 07/05/2017 15:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KENNEDY,DANIEL,L
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: KENNEDY,DANIEL
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:51

SUBJECT'S LAST NAME: FREIDES

DATE/TIME OF TEST: 06/05/2017 21:22:01

Diagnostics Check	OK	21:15
Air Blank	0.000	21:16
Control Test	0.080	21:16
Air Blank	0.000	21:17
Subject Sample # 1	0.270	21:17
Breath Volume	2.750	21:17
Air Blank	0.000	21:18
Air Blank	0.000	21:20
Subject Sample # 2	0.267	21:20
Breath Volume	2.453	21:20
Air Blank	0.000	21:21
Control Test	0.080	21:21
Air Blank	0.000	21:21
Diagnostics Check	OK	21:22

Data Download Date/Time: 07/05/2017 15:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PATTERSON,D,S
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: HORNBECK,D
ARRESTING OFFICER AGENCY: FL Dept of AG
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 10:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:00

SUBJECT'S LAST NAME: FIDURSKI

DATE/TIME OF TEST: 06/07/2017 11:05:38

Diagnostics Check	OK	10:59
Air Blank	0.000	11:00
Control Test	0.080	11:00
Air Blank	0.000	11:00
Subject Sample # 1	0.016	11:01
Breath Volume	1.781	11:01
Air Blank	0.000	11:01
Air Blank	0.000	11:03
Subject Sample # 2	0.017	11:04
Breath Volume	2.035	11:04
Air Blank	0.000	11:04
Control Test	0.081	11:05
Air Blank	0.000	11:05
Diagnostics Check	OK	11:05

Data Download Date/Time: 07/05/2017 15:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CURTSINGER,C,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: JACKSON,D
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: SANTANA

DATE/TIME OF TEST: 06/09/2017 01:58:21

Diagnostics Check		OK	01:46	
Air Blank		0.000	01:47	
Control Test		0.079	01:47	
Air Blank		0.000	01:47	
Subject Sample	# 1	0.204	01:51	Volume Not Met
Breath Volume		0.210	01:51	
Air Blank		0.000	01:51	
Air Blank		0.000	01:53	
Subject Sample	# 2	0.174	01:56	Volume Not Met
Breath Volume		0.281	01:56	
Air Blank		0.000	01:57	
Control Test		0.079	01:57	
Air Blank		0.000	01:58	
Diagnostics Check		OK	01:58	

Data Download Date/Time: 07/05/2017 15:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARPENTER,CHRISTOPHER,J
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: CARPENTER,CHRISTOPHER
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:53

SUBJECT'S LAST NAME: DAVIS-DEUTH

DATE/TIME OF TEST: 06/10/2017 00:24:56

Diagnostics Check	OK	00:18
Air Blank	0.000	00:19
Control Test	0.079	00:19
Air Blank	0.000	00:20
Subject Sample # 1	0.000	00:20
Breath Volume	2.816	00:20
Air Blank	0.000	00:21
Air Blank	0.000	00:23
Subject Sample # 2	0.000	00:23
Breath Volume	1.433	00:23
Air Blank	0.000	00:24
Control Test	0.079	00:24
Air Blank	0.000	00:24
Diagnostics Check	OK	00:24

Data Download Date/Time: 07/05/2017 15:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KENNEDY,DANIEL,L
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: KENNEDY,DANIEL
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:58

SUBJECT'S LAST NAME: TEBBE

DATE/TIME OF TEST: 06/10/2017 00:33:30

Diagnostics Check	OK	00:26
Air Blank	0.000	00:27
Control Test	0.079	00:27
Air Blank	0.000	00:28
Subject Sample # 1	0.122	00:28
Breath Volume	1.699	00:28
Air Blank	0.000	00:29
Air Blank	0.000	00:31
Subject Sample # 2	0.119	00:32
Breath Volume	1.820	00:32
Air Blank	0.000	00:32
Control Test	0.079	00:32
Air Blank	0.000	00:33
Diagnostics Check	OK	00:33

Data Download Date/Time: 07/05/2017 15:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KENNEDY,DANIEL,L
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: KENNEDY,DANIEL
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:16

SUBJECT'S LAST NAME: ESTRADA

DATE/TIME OF TEST: 06/10/2017 03:46:36

Diagnostics Check	OK	03:40
Air Blank	0.000	03:40
Control Test	0.079	03:41
Air Blank	0.000	03:41
Subject Sample # 1	0.000	03:42
Breath Volume	3.703	03:42
Air Blank	0.000	03:42
Air Blank	0.000	03:44
Subject Sample # 2	0.000	03:45
Breath Volume	3.785	03:45
Air Blank	0.000	03:45
Control Test	0.079	03:46
Air Blank	0.000	03:46
Diagnostics Check	OK	03:46

Data Download Date/Time: 07/05/2017 15:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KENNEDY,DANIEL,L
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: KENNEDY,DANIEL
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 17:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:11

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 06/10/2017 18:45:28

Diagnostics Check	OK	18:36
Air Blank	0.000	18:36
Control Test	0.080	18:37
Air Blank	0.000	18:37
Subject Sample # 1	0.287	18:40
Breath Volume	1.136	18:40
Air Blank	0.000	18:41
Air Blank	0.000	18:42
Subject Sample # 2	0.271	18:43
Breath Volume	1.863	18:43
Air Blank	0.000	18:44
Control Test	0.078	18:44
Air Blank	0.000	18:45
Diagnostics Check	OK	18:45

Data Download Date/Time: 07/05/2017 15:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VADNEY,MICHELE,M
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: CHAPMAN,M
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:38

SUBJECT'S LAST NAME: DUMAS

DATE/TIME OF TEST: 06/11/2017 04:10:41

Diagnostics Check	OK	04:04
Air Blank	0.000	04:04
Control Test	0.079	04:05
Air Blank	0.000	04:05
Subject Sample # 1	0.131	04:06
Breath Volume	2.359	04:06
Air Blank	0.000	04:06
Air Blank	0.000	04:08
Subject Sample # 2	0.137	04:09
Breath Volume	2.367	04:09
Air Blank	0.000	04:09
Control Test	0.079	04:10
Air Blank	0.000	04:10
Diagnostics Check	OK	04:10

Data Download Date/Time: 07/05/2017 15:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,KEVIN,J
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: MCCOY,D
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: VITORINO

DATE/TIME OF TEST: 06/11/2017 04:30:27

Diagnostics Check	OK	04:23
Air Blank	0.000	04:24
Control Test	0.080	04:24
Air Blank	0.000	04:25
Subject Sample # 1	0.151	04:25
Breath Volume	2.550	04:25
Air Blank	0.000	04:26
Air Blank	0.000	04:28
Subject Sample # 2	0.152	04:28
Breath Volume	2.136	04:28
Air Blank	0.000	04:29
Control Test	0.079	04:29
Air Blank	0.000	04:30
Diagnostics Check	OK	04:30

Data Download Date/Time: 07/05/2017 15:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DRISCOLL,JAMES,P
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: DRISCOLL,JAMES
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: CANALES

DATE/TIME OF TEST: 06/12/2017 01:59:26

Diagnostics Check	OK	01:50
Air Blank	0.000	01:50
Control Test	0.079	01:51
Air Blank	0.000	01:51
Subject Sample # 1	0.137	01:52
Breath Volume	1.738	01:52
Air Blank	0.000	01:52
Air Blank	0.000	01:54
Subject Sample # 2	0.110	01:55
Breath Volume	2.363	01:55
Air Blank	0.000	01:55
Air Blank	0.000	01:57
Subject Sample # 3	0.128	01:57
Breath Volume	3.007	01:57
Air Blank	0.000	01:58
Control Test	0.079	01:58
Air Blank	0.000	01:59
Diagnostics Check	OK	01:59

Data Download Date/Time: 07/05/2017 15:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CURTO,THEODORE,D
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: LEONI,STEPHAN
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:42

SUBJECT'S LAST NAME: LARGE

DATE/TIME OF TEST: 06/14/2017 02:15:45

Diagnostics Check	OK	02:09
Air Blank	0.000	02:09
Control Test	0.080	02:10
Air Blank	0.000	02:10
Subject Sample # 1	0.000	02:11
Breath Volume	3.285	02:11
Air Blank	0.000	02:11
Air Blank	0.000	02:13
Subject Sample # 2	0.000	02:14
Breath Volume	1.960	02:14
Air Blank	0.000	02:14
Control Test	0.079	02:15
Air Blank	0.000	02:15
Diagnostics Check	OK	02:15

Data Download Date/Time: 07/05/2017 15:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NATSON,ANDRAE,M
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: BRUGAL,N
ARRESTING OFFICER AGENCY: Naples PD
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 06/15/2017 00:42:22

Diagnostics Check	OK	00:36
Air Blank	0.000	00:36
Control Test	0.079	00:37
Air Blank	0.000	00:37
Subject Sample # 1	0.037	00:38
Breath Volume	2.164	00:38
Air Blank	0.000	00:38
Air Blank	0.000	00:40
Subject Sample # 2	0.034	00:40
Breath Volume	1.703	00:40
Air Blank	0.000	00:41
Control Test	0.080	00:41
Air Blank	0.000	00:42
Diagnostics Check	OK	00:42

Data Download Date/Time: 07/05/2017 15:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARPENTER,CHRISTOPHER,J
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: CARPENTER,CHRISTOPHER
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:16

SUBJECT'S LAST NAME: PATTERSON

DATE/TIME OF TEST: 06/15/2017 01:50:48

Diagnostics Check	OK	01:44
Air Blank	0.000	01:44
Control Test	0.080	01:44
Air Blank	0.000	01:45
Subject Sample # 1	0.137	01:46
Breath Volume	1.652	01:46
Air Blank	0.000	01:46
Air Blank	0.000	01:48
Subject Sample # 2	0.124	01:49
Breath Volume	2.042	01:49
Air Blank	0.000	01:49
Control Test	0.079	01:50
Air Blank	0.000	01:50
Diagnostics Check	OK	01:50

Data Download Date/Time: 07/05/2017 15:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARPENTER,CHRISTOPHER,J
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: CARPENTER,CHRISTOPHER
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:02

SUBJECT'S LAST NAME: NAGY

DATE/TIME OF TEST: 06/16/2017 01:32:56

Diagnostics Check	OK	01:25
Air Blank	0.000	01:26
Control Test	0.079	01:26
Air Blank	0.000	01:26
Subject Sample # 1	0.220	01:28
Breath Volume	3.644	01:28
Air Blank	0.000	01:28
Air Blank	0.000	01:30
Subject Sample # 2	0.216	01:31
Breath Volume	1.976	01:31
Air Blank	0.000	01:32
Control Test	0.079	01:32
Air Blank	0.000	01:32
Diagnostics Check	OK	01:32

Data Download Date/Time: 07/05/2017 15:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,ONDREA,L
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: STAUFFER,JASON
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:34

SUBJECT'S LAST NAME: FERRELL

DATE/TIME OF TEST: 06/17/2017 02:04:30

Diagnostics Check		OK	01:58
Air Blank		0.000	01:58
Control Test		0.080	01:58
Air Blank		0.000	01:59
Subject Sample	# 1	0.239	02:00
Breath Volume		1.179	02:00
Air Blank		0.000	02:00
Air Blank		0.000	02:02
Subject Sample	# 2	0.222	02:02
Breath Volume		1.718	02:02
Air Blank		0.000	02:03
Control Test		0.079	02:03
Air Blank		0.000	02:04
Diagnostics Check		OK	02:04

Data Download Date/Time: 07/05/2017 15:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARKS,D,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: STAUFFER,J
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 18:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:50

SUBJECT'S LAST NAME: WAHL

DATE/TIME OF TEST: 06/17/2017 20:22:45

Diagnostics Check	OK	20:16
Air Blank	0.000	20:17
Control Test	0.080	20:17
Air Blank	0.000	20:17
Subject Sample # 1	0.222	20:18
Breath Volume	2.582	20:18
Air Blank	0.000	20:19
Air Blank	0.000	20:20
Subject Sample # 2	0.224	20:21
Breath Volume	2.847	20:21
Air Blank	0.000	20:21
Control Test	0.079	20:22
Air Blank	0.000	20:22
Diagnostics Check	OK	20:22

Data Download Date/Time: 07/05/2017 15:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CURTO,THEODORE,D
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: CALDERON,A
ARRESTING OFFICER AGENCY: Naples PD
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:49

SUBJECT'S LAST NAME: RESENDIZ

DATE/TIME OF TEST: 06/18/2017 04:22:11

Diagnostics Check	OK	04:13	
Air Blank	0.000	04:14	
Control Test	0.079	04:14	
Air Blank	0.000	04:14	
Subject Sample	# 1 0.111	04:18	Volume Not Met
Breath Volume	0.367	04:18	
Air Blank	0.000	04:18	
Air Blank	0.000	04:20	
Subject Sample	# 2	04:20	Subject Test Refused
Breath Volume	0.000	04:20	
Air Blank	0.000	04:21	
Control Test	0.079	04:21	
Air Blank	0.000	04:22	
Diagnostics Check	OK	04:22	

Data Download Date/Time: 07/05/2017 15:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DRISCOLL,JAMES,P
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: DRISCOLL,JAMES
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 18:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:18

SUBJECT'S LAST NAME: KUHRE

DATE/TIME OF TEST: 06/20/2017 19:47:25

Diagnostics Check	OK	19:40
Air Blank	0.000	19:41
Control Test	0.080	19:41
Air Blank	0.000	19:42
Subject Sample # 1	0.100	19:42
Breath Volume	3.265	19:42
Air Blank	0.000	19:43
Air Blank	0.000	19:45
Subject Sample # 2	0.096	19:45
Breath Volume	3.105	19:45
Air Blank	0.000	19:46
Control Test	0.079	19:46
Air Blank	0.000	19:47
Diagnostics Check	OK	19:47

Data Download Date/Time: 07/05/2017 15:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DECKER,STEPHEN,J
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: ALLEN,R
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: WATSON

DATE/TIME OF TEST: 06/21/2017 00:47:10

Diagnostics Check	OK	00:41
Air Blank	0.000	00:41
Control Test	0.079	00:42
Air Blank	0.000	00:42
Subject Sample # 1	0.161	00:42
Breath Volume	1.804	00:42
Air Blank	0.000	00:43
Air Blank	0.000	00:45
Subject Sample # 2	0.156	00:45
Breath Volume	1.859	00:45
Air Blank	0.000	00:46
Control Test	0.079	00:46
Air Blank	0.000	00:47
Diagnostics Check	OK	00:47

Data Download Date/Time: 07/05/2017 15:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIERMAN,BRENT,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: COLLINS,J
ARRESTING OFFICER AGENCY: Naples PD
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:32

SUBJECT'S LAST NAME: TORMEY

DATE/TIME OF TEST: 06/21/2017 21:07:41

Diagnostics Check		OK	20:57	
Air Blank		0.000	20:58	
Control Test		0.078	20:58	
Air Blank		0.000	20:58	
Subject Sample	# 1	0.224	21:01	Volume Not Met
Breath Volume		0.207	21:01	
Air Blank		0.000	21:02	
Air Blank		0.000	21:04	
Subject Sample	# 2		21:06	Subject Test Refused
Breath Volume		0.144	21:06	
Air Blank		0.000	21:06	
Control Test		0.077	21:07	
Air Blank		0.000	21:07	
Diagnostics Check		OK	21:07	

Data Download Date/Time: 07/05/2017 15:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KENNEDY,DANIEL,L
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: KENNEDY,DANIEL
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:22

SUBJECT'S LAST NAME: MANGA

DATE/TIME OF TEST: 06/24/2017 02:01:09

Diagnostics Check	OK	01:47		
Air Blank	0.000	01:48		
Control Test	0.079	01:48		
Air Blank	0.000	01:49		
Subject Sample	# 1	01:52	No Sample Provided	
Breath Volume	0.000	01:52		
Air Blank	0.000	01:52		
Air Blank	0.000	01:54		
Subject Sample	# 2	0.216	01:55	No .020 agreement
Breath Volume	1.863	01:55		
Air Blank	0.000	01:56		
Air Blank	0.000	01:58		
Subject Sample	# 3	0.195	01:59	No .020 agreement
Breath Volume	1.550	01:59		
Air Blank	0.000	02:00		
Control Test	0.080	02:00		
Air Blank	0.000	02:01		
Diagnostics Check	OK	02:01		

Data Download Date/Time: 07/05/2017 15:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARPENTER,CHRISTOPHER,J
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: CARPENTER,CHRISTOPHER
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:54

SUBJECT'S LAST NAME: PADILLA

DATE/TIME OF TEST: 06/24/2017 03:26:45

Diagnostics Check	OK	03:18
Air Blank	0.000	03:18
Control Test	0.079	03:19
Air Blank	0.000	03:19
Subject Sample # 1	0.188	03:22
Breath Volume	1.808	03:22
Air Blank	0.000	03:22
Air Blank	0.000	03:24
Subject Sample # 2	0.202	03:25
Breath Volume	1.785	03:25
Air Blank	0.000	03:25
Control Test	0.078	03:26
Air Blank	0.000	03:26
Diagnostics Check	OK	03:26

Data Download Date/Time: 07/05/2017 15:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ASHLINE,C,J
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: YUREWITCH,K
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:19

SUBJECT'S LAST NAME: METZLER

DATE/TIME OF TEST: 06/24/2017 23:50:38

Diagnostics Check	OK	23:44
Air Blank	0.000	23:44
Control Test	0.079	23:45
Air Blank	0.000	23:45
Subject Sample # 1	0.109	23:46
Breath Volume	1.785	23:46
Air Blank	0.000	23:46
Air Blank	0.000	23:48
Subject Sample # 2	0.105	23:49
Breath Volume	1.562	23:49
Air Blank	0.000	23:49
Control Test	0.080	23:50
Air Blank	0.000	23:50
Diagnostics Check	OK	23:50

Data Download Date/Time: 07/05/2017 15:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORR,C,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: DEROSA,A
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:32

SUBJECT'S LAST NAME: JUAN JAIMES

DATE/TIME OF TEST: 06/25/2017 03:06:40

Diagnostics Check	OK	02:59
Air Blank	0.000	02:59
Control Test	0.079	03:00
Air Blank	0.000	03:00
Subject Sample # 1	0.213	03:01
Breath Volume	2.414	03:01
Air Blank	0.000	03:02
Air Blank	0.000	03:04
Subject Sample # 2	0.225	03:05
Breath Volume	2.933	03:05
Air Blank	0.000	03:05
Control Test	0.079	03:06
Air Blank	0.000	03:06
Diagnostics Check	OK	03:06

Data Download Date/Time: 07/05/2017 15:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KENNEDY,DANIEL,L
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: KENNEDY,DANIEL
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:18

SUBJECT'S LAST NAME: GRAHAM

DATE/TIME OF TEST: 06/25/2017 03:49:07

Diagnostics Check	OK	03:42
Air Blank	0.000	03:43
Control Test	0.080	03:43
Air Blank	0.000	03:44
Subject Sample # 1	0.159	03:44
Breath Volume	4.011	03:44
Air Blank	0.000	03:45
Air Blank	0.000	03:47
Subject Sample # 2	0.157	03:47
Breath Volume	4.027	03:47
Air Blank	0.000	03:48
Control Test	0.080	03:48
Air Blank	0.000	03:49
Diagnostics Check	OK	03:49

Data Download Date/Time: 07/05/2017 15:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORNELIUS,TIM,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: CORNELIUS,TIM
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:15

SUBJECT'S LAST NAME: VELASCO MORALES

DATE/TIME OF TEST: 06/25/2017 04:47:38

Diagnostics Check	OK	04:41
Air Blank	0.000	04:41
Control Test	0.079	04:41
Air Blank	0.000	04:42
Subject Sample # 1	0.156	04:43
Breath Volume	2.144	04:43
Air Blank	0.000	04:43
Air Blank	0.000	04:45
Subject Sample # 2	0.153	04:46
Breath Volume	2.644	04:46
Air Blank	0.000	04:46
Control Test	0.079	04:47
Air Blank	0.000	04:47
Diagnostics Check	OK	04:47

Data Download Date/Time: 07/05/2017 15:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KENNEDY,DANIEL,L
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: KENNEDY,DANIEL
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 17:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:19

SUBJECT'S LAST NAME: LEWIS JR

DATE/TIME OF TEST: 06/25/2017 18:55:15

Diagnostics Check	OK	18:49
Air Blank	0.000	18:49
Control Test	0.079	18:49
Air Blank	0.000	18:50
Subject Sample # 1	0.190	18:50
Breath Volume	2.558	18:50
Air Blank	0.000	18:51
Air Blank	0.000	18:53
Subject Sample # 2	0.190	18:53
Breath Volume	2.292	18:53
Air Blank	0.000	18:54
Control Test	0.079	18:54
Air Blank	0.000	18:55
Diagnostics Check	OK	18:55

Data Download Date/Time: 07/05/2017 15:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NOEL,Y,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: KLEIS,K
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: CASTILLO

DATE/TIME OF TEST: 06/26/2017 00:21:58

Diagnostics Check	OK	00:15
Air Blank	0.000	00:16
Control Test	0.079	00:16
Air Blank	0.000	00:17
Subject Sample # 1	0.239	00:17
Breath Volume	3.074	00:17
Air Blank	0.000	00:18
Air Blank	0.000	00:20
Subject Sample # 2	0.227	00:20
Breath Volume	2.113	00:20
Air Blank	0.000	00:21
Control Test	0.079	00:21
Air Blank	0.000	00:21
Diagnostics Check	OK	00:21

Data Download Date/Time: 07/05/2017 15:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROGERS, MATTHEW, L
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: SACKMANN, A
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 18:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:02

SUBJECT'S LAST NAME: ARANA

DATE/TIME OF TEST: 06/26/2017 20:36:11

Diagnostics Check	OK	20:30
Air Blank	0.000	20:30
Control Test	0.079	20:30
Air Blank	0.000	20:31
Subject Sample # 1	0.013	20:31
Breath Volume	2.457	20:31
Air Blank	0.000	20:32
Air Blank	0.000	20:34
Subject Sample # 2	0.013	20:34
Breath Volume	2.257	20:34
Air Blank	0.000	20:35
Control Test	0.080	20:35
Air Blank	0.000	20:36
Diagnostics Check	OK	20:36

Data Download Date/Time: 07/05/2017 15:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIBEY,AMANDA,L
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: MERCADO,D
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: CLOPTON

DATE/TIME OF TEST: 06/28/2017 00:14:01

Diagnostics Check		OK	00:00	
Air Blank		0.000	00:00	
Control Test		0.079	00:00	
Air Blank		0.000	00:01	
Subject Sample	# 1	0.135	00:04	Volume Not Met
Breath Volume		0.156	00:04	
Air Blank		0.000	00:05	
Air Blank		0.000	00:07	
Subject Sample	# 2	0.150	00:09	
Breath Volume		2.347	00:09	
Air Blank		0.000	00:09	
Air Blank		0.000	00:11	
Subject Sample	# 3	0.148	00:12	
Breath Volume		2.218	00:12	
Air Blank		0.000	00:13	
Control Test		0.079	00:13	
Air Blank		0.000	00:13	
Diagnostics Check		OK	00:14	

Data Download Date/Time: 07/05/2017 15:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,JENNIFER,S
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: ROWLEY,C
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 10:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 11:28

SUBJECT'S LAST NAME: PIETERSE

DATE/TIME OF TEST: 06/28/2017 11:59:39

Diagnostics Check		OK	11:53
Air Blank		0.000	11:53
Control Test		0.079	11:53
Air Blank		0.000	11:54
Subject Sample	# 1	0.163	11:55
Breath Volume		3.832	11:55
Air Blank		0.000	11:55
Air Blank		0.000	11:57
Subject Sample	# 2	0.143	11:58
Breath Volume		2.828	11:58
Air Blank		0.000	11:58
Control Test		0.079	11:59
Air Blank		0.000	11:59
Diagnostics Check		OK	11:59

Data Download Date/Time: 07/05/2017 15:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALLOPS,COLTON,M
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: MILLER,K
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:27

SUBJECT'S LAST NAME: ALMANZA

DATE/TIME OF TEST: 06/29/2017 04:00:09

Diagnostics Check	OK	03:53
Air Blank	0.000	03:54
Control Test	0.080	03:54
Air Blank	0.000	03:54
Subject Sample # 1	0.148	03:55
Breath Volume	2.425	03:55
Air Blank	0.000	03:56
Air Blank	0.000	03:58
Subject Sample # 2	0.149	03:58
Breath Volume	1.125	03:58
Air Blank	0.000	03:59
Control Test	0.079	03:59
Air Blank	0.000	04:00
Diagnostics Check	OK	04:00

Data Download Date/Time: 07/05/2017 15:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARGO,TIMOTHY,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: WARGO,TIMOTHY
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 05/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: CARTER

DATE/TIME OF TEST: 06/01/2017 02:56:40

Diagnostics Check	OK	02:50
Air Blank	0.000	02:50
Control Test	0.078	02:50
Air Blank	0.000	02:51
Subject Sample # 1	0.151	02:52
Breath Volume	1.843	02:52
Air Blank	0.000	02:52
Air Blank	0.000	02:54
Subject Sample # 2	0.157	02:55
Breath Volume	1.625	02:55
Air Blank	0.000	02:55
Control Test	0.079	02:56
Air Blank	0.000	02:56
Diagnostics Check	OK	02:56

Data Download Date/Time: 06/02/2017 15:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KENNEDY,DANIEL,L
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: KENNEDY,DANIEL
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 05/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: KOKA

DATE/TIME OF TEST: 06/01/2017 03:38:27

Diagnostics Check	OK	03:32
Air Blank	0.000	03:32
Control Test	0.078	03:33
Air Blank	0.000	03:33
Subject Sample # 1	0.158	03:34
Breath Volume	2.933	03:34
Air Blank	0.000	03:34
Air Blank	0.000	03:36
Subject Sample # 2	0.159	03:36
Breath Volume	1.160	03:36
Air Blank	0.000	03:37
Control Test	0.078	03:37
Air Blank	0.000	03:38
Diagnostics Check	OK	03:38

Data Download Date/Time: 06/02/2017 15:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROHRAFF,HARRY,E
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: GINTHER,D
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 05/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: MOORE

DATE/TIME OF TEST: 06/02/2017 01:25:39

Diagnostics Check	OK	01:18
Air Blank	0.000	01:19
Control Test	0.079	01:19
Air Blank	0.000	01:20
Subject Sample # 1	0.085	01:20
Breath Volume	2.332	01:20
Air Blank	0.000	01:21
Air Blank	0.000	01:23
Subject Sample # 2	0.080	01:24
Breath Volume	1.460	01:24
Air Blank	0.000	01:24
Control Test	0.078	01:25
Air Blank	0.000	01:25
Diagnostics Check	OK	01:25

Data Download Date/Time: 06/02/2017 15:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000947

INSTRUMENT REGISTERED TO: GLADES COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONDRAGON, EDWIN, F
BREATH TEST OPERATOR AGENCY: Glades County SO
ARRESTING OFFICER: HOWELL, KEVIN
ARRESTING OFFICER AGENCY: Glades County SO
LAST AGENCY INSPECTION: 05/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04216080A3
DRY GAS EXPIRY DATE: 03/05/2018
VIOLATION CODE: ADMIN
ARREST TIME: 09:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:49

SUBJECT'S LAST NAME: PEREZ MIRANDA

DATE/TIME OF TEST: 06/11/2017 10:33:15

Diagnostics Check	OK	10:23
Air Blank	0.000	10:23
Control Test	0.077	10:23
Air Blank	0.000	10:24
Subject Sample # 1	0.296	10:25
Breath Volume	1.437	10:25
Air Blank	0.000	10:26
Air Blank	0.000	10:28
Subject Sample # 2	0.319	10:28
Breath Volume	1.816	10:28
Air Blank	0.000	10:29
Air Blank	0.000	10:30
Subject Sample # 3	0.299	10:31
Breath Volume	1.824	10:31
Air Blank	0.000	10:32
Control Test	0.076	10:32
Air Blank	0.000	10:33
Diagnostics Check	OK	10:33

Data Download Date/Time: 06/30/2017 15:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000947

INSTRUMENT REGISTERED TO: GLADES COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MADRUGO,FELICIA,
BREATH TEST OPERATOR AGENCY: Glades County SO
ARRESTING OFFICER: SOTO,MICHAEL
ARRESTING OFFICER AGENCY: Glades County SO
LAST AGENCY INSPECTION: 05/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04216080A3
DRY GAS EXPIRY DATE: 03/05/2018
VIOLATION CODE: OTHER
ARREST TIME: 01:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: SHAW

DATE/TIME OF TEST: 06/17/2017 02:03:34

Diagnostics Check		OK	01:57
Air Blank		0.000	01:57
Control Test		0.077	01:58
Air Blank		0.000	01:58
Subject Sample	# 1	0.123	01:59
Breath Volume		2.707	01:59
Air Blank		0.000	01:59
Air Blank		0.000	02:01
Subject Sample	# 2	0.118	02:01
Breath Volume		2.558	02:01
Air Blank		0.000	02:02
Control Test		0.077	02:02
Air Blank		0.000	02:03
Diagnostics Check		OK	02:03

Data Download Date/Time: 06/30/2017 15:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000947

INSTRUMENT REGISTERED TO: GLADES COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HICKS,CHRISTINA,L
BREATH TEST OPERATOR AGENCY: Glades County SO
ARRESTING OFFICER: SOTO,MICHAEL
ARRESTING OFFICER AGENCY: Glades County SO
LAST AGENCY INSPECTION: 05/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04216080A3
DRY GAS EXPIRY DATE: 03/05/2018
VIOLATION CODE: ADMIN
ARREST TIME: 00:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:30

SUBJECT'S LAST NAME: SHAW

DATE/TIME OF TEST: 06/17/2017 08:19:14

Diagnostics Check	OK	08:13
Air Blank	0.000	08:13
Control Test	0.078	08:14
Air Blank	0.000	08:14
Subject Sample # 1	0.021	08:14
Breath Volume	1.425	08:14
Air Blank	0.000	08:15
Air Blank	0.000	08:17
Subject Sample # 2	0.019	08:17
Breath Volume	1.207	08:17
Air Blank	0.000	08:18
Control Test	0.076	08:18
Air Blank	0.000	08:19
Diagnostics Check	OK	08:19

Data Download Date/Time: 06/30/2017 15:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000947

INSTRUMENT REGISTERED TO: GLADES COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MADRUGO,FELICIA,
BREATH TEST OPERATOR AGENCY: Glades County SO
ARRESTING OFFICER: CLENDENING,JAMES
ARRESTING OFFICER AGENCY: Glades County SO
LAST AGENCY INSPECTION: 05/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04216080A3
DRY GAS EXPIRY DATE: 03/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:04

SUBJECT'S LAST NAME: BEAVER

DATE/TIME OF TEST: 06/19/2017 00:41:03

Diagnostics Check	OK	00:25	
Air Blank	0.000	00:25	
Control Test	0.078	00:26	
Air Blank	0.000	00:26	
Subject Sample	# 1 0.160	00:29	Volume Not Met
Breath Volume	0.050	00:29	
Air Blank	0.000	00:30	
Air Blank	0.000	00:32	
Subject Sample	# 2 0.195	00:34	
Breath Volume	1.246	00:34	
Air Blank	0.000	00:35	
Air Blank	0.000	00:37	
Subject Sample	# 3 0.191	00:39	
Breath Volume	1.667	00:39	
Air Blank	0.000	00:40	
Control Test	0.073	00:40	Control Outside Tolerance
Air Blank	0.000	00:41	

Data Download Date/Time: 06/30/2017 15:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000947

INSTRUMENT REGISTERED TO: GLADES COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BONILLA, JOSUE, D
BREATH TEST OPERATOR AGENCY: Glades County SO
ARRESTING OFFICER: PORTTORFF, B
ARRESTING OFFICER AGENCY: Seminole Dept of LE
LAST AGENCY INSPECTION: 05/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04216080A3
DRY GAS EXPIRY DATE: 03/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:11

SUBJECT'S LAST NAME: GOPHER

DATE/TIME OF TEST: 06/22/2017 02:47:44

Diagnostics Check	OK	02:43	
Air Blank	0.000	02:44	
Control Test	0.078	02:44	
Air Blank	0.000	02:44	
Subject Sample	# 1	02:46	Subject Test Refused
Breath Volume	0.000	02:46	
Air Blank	0.000	02:46	
Control Test	0.078	02:47	
Air Blank	0.000	02:47	
Diagnostics Check	OK	02:47	

Data Download Date/Time: 06/30/2017 15:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000947

INSTRUMENT REGISTERED TO: GLADES COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLENDENING,JAMES,T
BREATH TEST OPERATOR AGENCY: Glades County SO
ARRESTING OFFICER: CLENDENING,JAMES
ARRESTING OFFICER AGENCY: Glades County SO
LAST AGENCY INSPECTION: 05/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04216080A3
DRY GAS EXPIRY DATE: 03/05/2018
VIOLATION CODE: DUI
ARREST TIME: 17:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:00

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 06/26/2017 18:37:12

Diagnostics Check	OK	18:31
Air Blank	0.000	18:31
Control Test	0.077	18:31
Air Blank	0.000	18:32
Subject Sample # 1	0.215	18:32
Breath Volume	1.734	18:32
Air Blank	0.000	18:33
Air Blank	0.000	18:35
Subject Sample # 2	0.224	18:35
Breath Volume	2.847	18:35
Air Blank	0.000	18:36
Control Test	0.076	18:36
Air Blank	0.000	18:37
Diagnostics Check	OK	18:37

Data Download Date/Time: 06/30/2017 15:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000948

INSTRUMENT REGISTERED TO: GLADES COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BONILLA, JOSUE, D
BREATH TEST OPERATOR AGENCY: Glades County SO
ARRESTING OFFICER: MCQUIAD, D
ARRESTING OFFICER AGENCY: Glades County SO
LAST AGENCY INSPECTION: 05/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04216080A3
DRY GAS EXPIRY DATE: 03/05/2018
VIOLATION CODE: ADMIN
ARREST TIME: 20:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:20

SUBJECT'S LAST NAME: CLINE

DATE/TIME OF TEST: 06/04/2017 20:54:19

Diagnostics Check	OK	20:48
Air Blank	0.000	20:48
Control Test	0.077	20:48
Air Blank	0.000	20:49
Subject Sample # 1	0.251	20:49
Breath Volume	3.535	20:49
Air Blank	0.000	20:50
Air Blank	0.000	20:52
Subject Sample # 2	0.263	20:52
Breath Volume	3.417	20:52
Air Blank	0.000	20:53
Control Test	0.076	20:53
Air Blank	0.000	20:54
Diagnostics Check	OK	20:54

Data Download Date/Time: 06/30/2017 15:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000948

INSTRUMENT REGISTERED TO: GLADES COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BONILLA, JOSUE, D
BREATH TEST OPERATOR AGENCY: Glades County SO
ARRESTING OFFICER: MCQUAID, D
ARRESTING OFFICER AGENCY: Glades County SO
LAST AGENCY INSPECTION: 05/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04216080A3
DRY GAS EXPIRY DATE: 03/05/2018
VIOLATION CODE: ADMIN
ARREST TIME: 20:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: CLINE

DATE/TIME OF TEST: 06/05/2017 03:54:48

Diagnostics Check	OK	03:48
Air Blank	0.000	03:48
Control Test	0.078	03:48
Air Blank	0.000	03:49
Subject Sample # 1	0.147	03:50
Breath Volume	3.316	03:50
Air Blank	0.000	03:50
Air Blank	0.000	03:52
Subject Sample # 2	0.157	03:53
Breath Volume	2.945	03:53
Air Blank	0.000	03:53
Control Test	0.077	03:54
Air Blank	0.000	03:54
Diagnostics Check	OK	03:54

Data Download Date/Time: 06/30/2017 15:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000948

INSTRUMENT REGISTERED TO: GLADES COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAXTER,AUDREY,D
BREATH TEST OPERATOR AGENCY: Glades County SO
ARRESTING OFFICER: MCQUAID,DWAYNE
ARRESTING OFFICER AGENCY: Glades County SO
LAST AGENCY INSPECTION: 05/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04216080A3
DRY GAS EXPIRY DATE: 03/05/2018
VIOLATION CODE: ADMIN
ARREST TIME: 08:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:02

SUBJECT'S LAST NAME: CLINE

DATE/TIME OF TEST: 06/05/2017 06:39:18

Diagnostics Check	OK	06:32
Air Blank	0.000	06:32
Control Test	0.078	06:32
Air Blank	0.000	06:33
Subject Sample # 1	0.096	06:33
Breath Volume	1.488	06:33
Air Blank	0.000	06:34
Air Blank	0.000	06:36
Subject Sample # 2	0.089	06:37
Breath Volume	2.015	06:37
Air Blank	0.000	06:38
Control Test	0.078	06:38
Air Blank	0.000	06:39
Diagnostics Check	OK	06:39

Data Download Date/Time: 06/30/2017 15:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000948

INSTRUMENT REGISTERED TO: GLADES COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAXTER,AUDREY,D
BREATH TEST OPERATOR AGENCY: Glades County SO
ARRESTING OFFICER: MCQUAID,DWAYNE
ARRESTING OFFICER AGENCY: Glades County SO
LAST AGENCY INSPECTION: 05/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04216080A3
DRY GAS EXPIRY DATE: 03/05/2018
VIOLATION CODE: ADMIN
ARREST TIME: 08:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:42

SUBJECT'S LAST NAME: CLINE

DATE/TIME OF TEST: 06/05/2017 08:12:33

Diagnostics Check	OK	08:06
Air Blank	0.000	08:06
Control Test	0.079	08:07
Air Blank	0.000	08:07
Subject Sample # 1	0.062	08:08
Breath Volume	1.238	08:08
Air Blank	0.000	08:08
Air Blank	0.000	08:10
Subject Sample # 2	0.057	08:10
Breath Volume	1.195	08:10
Air Blank	0.000	08:11
Control Test	0.077	08:11
Air Blank	0.000	08:12
Diagnostics Check	OK	08:12

Data Download Date/Time: 06/30/2017 15:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000950

INSTRUMENT REGISTERED TO: HAVANA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FINUFF,DANIEL,W
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: FINUFF,DANIEL
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 05/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 05:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:22

SUBJECT'S LAST NAME: ABRIGO

DATE/TIME OF TEST: 06/04/2017 05:46:21

Diagnostics Check	OK	05:43	
Air Blank	0.000	05:43	
Control Test	0.076	05:44	
Air Blank	0.000	05:44	
Subject Sample	# 1	05:45	Subject Test Refused
Breath Volume	0.000	05:45	
Air Blank	0.000	05:45	
Control Test	0.077	05:45	
Air Blank	0.000	05:46	
Diagnostics Check	OK	05:46	

Data Download Date/Time: 06/21/2017 11:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000950

INSTRUMENT REGISTERED TO: HAVANA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEWIS,KENNETH,G
BREATH TEST OPERATOR AGENCY: Havana PD
ARRESTING OFFICER: LEWIS,KENNETH
ARRESTING OFFICER AGENCY: Havana PD
LAST AGENCY INSPECTION: 06/21/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:03

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 06/24/2017 02:33:25

Diagnostics Check	OK	02:25
Air Blank	0.000	02:26
Control Test	0.076	02:26
Air Blank	0.000	02:26
Subject Sample # 1	0.199	02:27
Breath Volume	1.886	02:27
Air Blank	0.000	02:28
Air Blank	0.000	02:30
Subject Sample # 2	0.196	02:31
Breath Volume	1.625	02:31
Air Blank	0.000	02:32
Control Test	0.075	02:32
Air Blank	0.000	02:33
Diagnostics Check	OK	02:33

Data Download Date/Time: 07/05/2017 10:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000951

INSTRUMENT REGISTERED TO: HENDRY COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NORWOOD,CHRISTOPHER,J
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: GREEN,CLIFTON
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 05/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 83551
DRY GAS EXPIRY DATE: 10/02/2018
VIOLATION CODE: DUI
ARREST TIME: 00:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:29

SUBJECT'S LAST NAME: TRUIZ

DATE/TIME OF TEST: 06/10/2017 01:12:10

Diagnostics Check	OK	01:09	
Air Blank	0.000	01:09	
Control Test	0.080	01:10	
Air Blank	0.000	01:10	
Subject Sample	# 1	01:10	Subject Test Refused
Breath Volume	0.000	01:10	
Air Blank	0.000	01:11	
Control Test	0.080	01:11	
Air Blank	0.000	01:12	
Diagnostics Check	OK	01:12	

Data Download Date/Time: 06/14/2017 03:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000951

INSTRUMENT REGISTERED TO: HENDRY COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NORWOOD,CHRISTOPHER,J
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: WOODS,JOSHUA
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 05/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 83551
DRY GAS EXPIRY DATE: 10/02/2018
VIOLATION CODE: DUI
ARREST TIME: 00:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: ACOSTA

DATE/TIME OF TEST: 06/10/2017 01:34:59

Diagnostics Check		OK	01:28
Air Blank		0.000	01:29
Control Test		0.080	01:29
Air Blank		0.000	01:30
Subject Sample	# 1	0.181	01:30
Breath Volume		3.562	01:30
Air Blank		0.000	01:31
Air Blank		0.000	01:33
Subject Sample	# 2	0.186	01:33
Breath Volume		3.464	01:33
Air Blank		0.000	01:34
Control Test		0.081	01:34
Air Blank		0.000	01:34
Diagnostics Check		OK	01:34

Data Download Date/Time: 06/14/2017 03:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000951

INSTRUMENT REGISTERED TO: HENDRY COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NORWOOD,CHRISTOPHER,J
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: NORWOOD,CHRISTOPHER
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 05/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 83551
DRY GAS EXPIRY DATE: 10/02/2018
VIOLATION CODE: DUI
ARREST TIME: 03:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:12

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 06/10/2017 03:47:56

Diagnostics Check	OK	03:41
Air Blank	0.000	03:41
Control Test	0.080	03:42
Air Blank	0.000	03:42
Subject Sample # 1	0.057	03:43
Breath Volume	2.234	03:43
Air Blank	0.000	03:43
Air Blank	0.000	03:45
Subject Sample # 2	0.051	03:46
Breath Volume	1.503	03:46
Air Blank	0.000	03:47
Control Test	0.082	03:47
Air Blank	0.000	03:47
Diagnostics Check	OK	03:47

Data Download Date/Time: 06/14/2017 03:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000952

INSTRUMENT REGISTERED TO: FRANKLIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBB,THOMAS,J
BREATH TEST OPERATOR AGENCY: Franklin County SO
ARRESTING OFFICER: HAMM,JAMES
ARRESTING OFFICER AGENCY: Franklin County SO
LAST AGENCY INSPECTION: 05/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1841081
DRY GAS EXPIRY DATE: 04/08/2018
VIOLATION CODE: DUI
ARREST TIME: 20:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:36

SUBJECT'S LAST NAME: CASE

DATE/TIME OF TEST: 06/06/2017 21:05:51

Diagnostics Check	OK	20:59
Air Blank	0.000	21:00
Control Test	0.081	21:00
Air Blank	0.000	21:00
Subject Sample # 1	0.202	21:01
Breath Volume	1.886	21:01
Air Blank	0.000	21:01
Air Blank	0.000	21:03
Subject Sample # 2	0.202	21:04
Breath Volume	1.781	21:04
Air Blank	0.000	21:04
Control Test	0.078	21:05
Air Blank	0.000	21:05
Diagnostics Check	OK	21:05

Data Download Date/Time: 06/30/2017 10:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000952

INSTRUMENT REGISTERED TO: FRANKLIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBB,THOMAS,J
BREATH TEST OPERATOR AGENCY: Franklin County SO
ARRESTING OFFICER: HAMM,JAMES
ARRESTING OFFICER AGENCY: Franklin County SO
LAST AGENCY INSPECTION: 05/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1841081
DRY GAS EXPIRY DATE: 04/08/2018
VIOLATION CODE: DUI
ARREST TIME: 23:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:14

SUBJECT'S LAST NAME: BALLENGER

DATE/TIME OF TEST: 06/14/2017 23:50:40

Diagnostics Check		OK	23:34	
Air Blank		0.000	23:34	
Control Test		0.080	23:35	
Air Blank		0.000	23:35	
Subject Sample	# 1	0.154	23:38	Volume Not Met
Breath Volume		0.085	23:38	
Air Blank		0.000	23:39	
Air Blank		0.000	23:41	
Subject Sample	# 2	0.156	23:43	
Breath Volume		2.394	23:43	
Air Blank		0.000	23:44	
Air Blank		0.000	23:45	
Subject Sample	# 3	0.135	23:48	Volume Not Met
Breath Volume		0.812	23:48	
Air Blank		0.000	23:49	
Control Test		0.076	23:50	
Air Blank		0.000	23:50	
Diagnostics Check		OK	23:50	

Data Download Date/Time: 06/30/2017 10:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000952

INSTRUMENT REGISTERED TO: FRANKLIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHALEY,CARLTON,L
BREATH TEST OPERATOR AGENCY: Franklin County SO
ARRESTING OFFICER: CORTOPASSI,BARON
ARRESTING OFFICER AGENCY: Franklin County SO
LAST AGENCY INSPECTION: 05/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1841081
DRY GAS EXPIRY DATE: 04/08/2018
VIOLATION CODE: DUI
ARREST TIME: 20:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:15

SUBJECT'S LAST NAME: STRIPLING

DATE/TIME OF TEST: 06/16/2017 21:40:39

Diagnostics Check	OK	21:35	
Air Blank	0.000	21:36	
Control Test	0.080	21:36	
Air Blank	0.000	21:37	
Subject Sample	# 1 0.298	21:39	
Breath Volume	1.609	21:39	
Air Blank		21:39	Purge Fail
Air Blank	0.000	21:40	

Data Download Date/Time: 06/30/2017 10:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000952

INSTRUMENT REGISTERED TO: FRANKLIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHALEY,CARLTON,L
BREATH TEST OPERATOR AGENCY: Franklin County SO
ARRESTING OFFICER: CORTOPASSI,BARON
ARRESTING OFFICER AGENCY: Franklin County SO
LAST AGENCY INSPECTION: 05/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1841081
DRY GAS EXPIRY DATE: 04/08/2018
VIOLATION CODE: DUI
ARREST TIME: 20:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:15

SUBJECT'S LAST NAME: STRIPLING

DATE/TIME OF TEST: 06/16/2017 21:49:31

Diagnostics Check	OK	21:43
Air Blank	0.000	21:43
Control Test	0.078	21:43
Air Blank	0.000	21:44
Subject Sample # 1	0.302	21:44
Breath Volume	1.601	21:44
Air Blank	0.000	21:45
Air Blank	0.000	21:47
Subject Sample # 2	0.287	21:47
Breath Volume	1.406	21:47
Air Blank	0.000	21:48
Control Test	0.080	21:48
Air Blank	0.000	21:49
Diagnostics Check	OK	21:49

Data Download Date/Time: 06/30/2017 10:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000952

INSTRUMENT REGISTERED TO: FRANKLIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBB,THOMAS,J
BREATH TEST OPERATOR AGENCY: Franklin County SO
ARRESTING OFFICER: JOCHIM,AMBER
ARRESTING OFFICER AGENCY: Carrabelle PD
LAST AGENCY INSPECTION: 05/24/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1841081
DRY GAS EXPIRY DATE: 04/08/2018
VIOLATION CODE: DUI
ARREST TIME: 12:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:49

SUBJECT'S LAST NAME: MCCORD

DATE/TIME OF TEST: 06/25/2017 01:35:03

Diagnostics Check	OK	01:28
Air Blank	0.000	01:28
Control Test	0.080	01:29
Air Blank	0.000	01:29
Subject Sample # 1	0.171	01:30
Breath Volume	2.363	01:30
Air Blank	0.000	01:31
Air Blank	0.000	01:32
Subject Sample # 2	0.165	01:33
Breath Volume	2.750	01:33
Air Blank	0.000	01:34
Control Test	0.077	01:34
Air Blank	0.000	01:34
Diagnostics Check	OK	01:35

Data Download Date/Time: 06/30/2017 10:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000952

INSTRUMENT REGISTERED TO: FRANKLIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHALEY,CARLTON,L
BREATH TEST OPERATOR AGENCY: Franklin County SO
ARRESTING OFFICER: HAMM,JAMES
ARRESTING OFFICER AGENCY: Franklin County SO
LAST AGENCY INSPECTION: 05/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1841081
DRY GAS EXPIRY DATE: 04/08/2018
VIOLATION CODE: DUI
ARREST TIME: 21:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:18

SUBJECT'S LAST NAME: ADAMS

DATE/TIME OF TEST: 06/29/2017 22:44:47

Diagnostics Check	OK	22:38
Air Blank	0.000	22:38
Control Test	0.081	22:38
Air Blank	0.000	22:39
Subject Sample # 1	0.181	22:40
Breath Volume	1.941	22:40
Air Blank	0.000	22:40
Air Blank	0.000	22:42
Subject Sample # 2	0.167	22:43
Breath Volume	2.671	22:43
Air Blank	0.000	22:43
Control Test	0.081	22:44
Air Blank	0.000	22:44
Diagnostics Check	OK	22:44

Data Download Date/Time: 06/30/2017 10:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000954

INSTRUMENT REGISTERED TO: JEFFERSON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBINSON,DARIN,L
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: MATTHEWS,DUSTIN
ARRESTING OFFICER AGENCY: Jefferson County SO
LAST AGENCY INSPECTION: 05/03/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: CROFFOOT

DATE/TIME OF TEST: 06/03/2017 04:10:47

Diagnostics Check	OK	04:07	
Air Blank	0.000	04:08	
Control Test	0.080	04:08	
Air Blank	0.000	04:09	
Subject Sample	# 1	04:09	Subject Test Refused
Breath Volume	0.000	04:09	
Air Blank	0.000	04:09	
Control Test	0.078	04:10	
Air Blank	0.000	04:10	
Diagnostics Check	OK	04:10	

Data Download Date/Time: 06/28/2017 10:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000954

INSTRUMENT REGISTERED TO: JEFFERSON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WRIGHT, JOSHUA, N
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: CAMPER, ALAN
ARRESTING OFFICER AGENCY: Jefferson County SO
LAST AGENCY INSPECTION: 05/03/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:05

SUBJECT'S LAST NAME: STRINGER

DATE/TIME OF TEST: 06/16/2017 23:32:18

Diagnostics Check	OK	23:25	
Air Blank	0.000	23:26	
Control Test	0.080	23:26	
Air Blank	0.000	23:27	
Subject Sample # 1	0.078	23:27	
Breath Volume	3.628	23:27	
Air Blank	0.000	23:28	
Air Blank	0.000	23:30	
Subject Sample # 2	0.074	23:30	
Breath Volume	2.714	23:30	
Air Blank	0.000	23:31	
Control Test	0.078	23:31	RFI Detect
Air Blank	0.000	23:32	

Data Download Date/Time: 06/28/2017 10:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000954

INSTRUMENT REGISTERED TO: JEFFERSON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARKER,JAMES,I
BREATH TEST OPERATOR AGENCY: Jefferson County SO
ARRESTING OFFICER: RIGDON,JERROD
ARRESTING OFFICER AGENCY: Jefferson County SO
LAST AGENCY INSPECTION: 05/03/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:16

SUBJECT'S LAST NAME: PARRISH

DATE/TIME OF TEST: 06/22/2017 23:38:30

Diagnostics Check	OK	23:36	
Air Blank	0.000	23:36	
Control Test	0.079	23:37	
Air Blank	0.000	23:37	
Subject Sample	# 1	23:38	Sequence Aborted
Breath Volume	0.000	23:38	
Air Blank	0.000	23:38	

Data Download Date/Time: 06/28/2017 10:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000955

INSTRUMENT REGISTERED TO: CLEWISTON P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HESTER,JAMES,R
BREATH TEST OPERATOR AGENCY: Clewiston PD
ARRESTING OFFICER: MCLAIN,KYLE
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 05/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:27

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 06/04/2017 02:50:25

Diagnostics Check	OK	02:47	
Air Blank	0.000	02:48	
Control Test	0.079	02:48	
Air Blank	0.000	02:48	
Subject Sample	# 1	02:49	Subject Test Refused
Breath Volume	0.000	02:49	
Air Blank	0.000	02:49	
Control Test	0.079	02:49	
Air Blank	0.000	02:50	
Diagnostics Check	OK	02:50	

Data Download Date/Time: 06/14/2017 03:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000955

INSTRUMENT REGISTERED TO: CLEWISTON P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HESTER,JAMES,R
BREATH TEST OPERATOR AGENCY: Clewiston PD
ARRESTING OFFICER: BLACKMAN,ANTHONY
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 05/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 03:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: MOORE

DATE/TIME OF TEST: 06/08/2017 04:18:46

Diagnostics Check	OK	04:15	
Air Blank	0.000	04:16	
Control Test	0.080	04:16	
Air Blank	0.000	04:16	
Subject Sample	# 1	04:17	Subject Test Refused
Breath Volume	0.000	04:17	
Air Blank	0.000	04:17	
Control Test	0.080	04:18	
Air Blank	0.000	04:18	
Diagnostics Check	OK	04:18	

Data Download Date/Time: 06/14/2017 03:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000955

INSTRUMENT REGISTERED TO: CLEWISTON P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LAROSA,RYAN,M

BREATH TEST OPERATOR AGENCY: Clewiston PD

ARRESTING OFFICER: LAROSA,RYAN

ARRESTING OFFICER AGENCY: Clewiston PD

LAST AGENCY INSPECTION: 05/24/2017

Date or Time Changed: No

DRY GAS STANDARD LOT #: 644476

DRY GAS EXPIRY DATE: 12/01/2019

VIOLATION CODE: DUI

ARREST TIME: 01:32

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: CHRISTIE

DATE/TIME OF TEST: 06/11/2017 02:30:14

Diagnostics Check	OK	02:23
Air Blank	0.000	02:23
Control Test	0.080	02:24
Air Blank	0.000	02:24
Subject Sample # 1	0.112	02:25
Breath Volume	2.230	02:25
Air Blank	0.000	02:26
Air Blank	0.000	02:28
Subject Sample # 2	0.108	02:28
Breath Volume	2.375	02:28
Air Blank	0.000	02:29
Control Test	0.080	02:29
Air Blank	0.000	02:30
Diagnostics Check	OK	02:30

Data Download Date/Time: 06/14/2017 03:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000955

INSTRUMENT REGISTERED TO: CLEWISTON P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HESTER,JAMES,R

BREATH TEST OPERATOR AGENCY: Clewiston PD

ARRESTING OFFICER: MCNEIL,DARRIN

ARRESTING OFFICER AGENCY: Clewiston PD

LAST AGENCY INSPECTION: 06/14/2017

Date or Time Changed: No

DRY GAS STANDARD LOT #: 644476

DRY GAS EXPIRY DATE: 12/01/2019

VIOLATION CODE: DUI

ARREST TIME: 02:35

GENDER: Female

OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: ROBERTS

DATE/TIME OF TEST: 06/18/2017 03:02:45

Diagnostics Check	OK	02:59	
Air Blank	0.000	03:00	
Control Test	0.079	03:00	
Air Blank	0.000	03:01	
Subject Sample	# 1	03:01	Subject Test Refused
Breath Volume	0.000	03:01	
Air Blank	0.000	03:01	
Control Test	0.079	03:02	
Air Blank	0.000	03:02	
Diagnostics Check	OK	03:02	

Data Download Date/Time: 06/28/2017 11:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000960

INSTRUMENT REGISTERED TO: LIBERTY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FINUFF,DANIEL,W
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: FINUFF,DANIEL
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 06/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: WALKER

DATE/TIME OF TEST: 06/21/2017 01:28:46

Diagnostics Check	OK	01:25	
Air Blank	0.000	01:26	
Control Test	0.083	01:26	
Air Blank	0.000	01:27	
Subject Sample	# 1	01:27	Subject Test Refused
Breath Volume	0.000	01:27	
Air Blank	0.000	01:27	
Control Test	0.083	01:28	
Air Blank	0.000	01:28	
Diagnostics Check	OK	01:28	

Data Download Date/Time: 07/19/2017 10:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000961

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CROSS, SR.,MICHAEL,S
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: CROSS, SR.,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 06/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:47

SUBJECT'S LAST NAME: STEGE

DATE/TIME OF TEST: 06/18/2017 22:47:54

Diagnostics Check	OK	22:36	
Air Blank	0.000	22:36	
Control Test	0.081	22:37	
Air Blank	0.000	22:37	
Subject Sample	# 1	22:40	No Sample Provided
Breath Volume	0.000	22:40	
Air Blank	0.000	22:41	
Air Blank	0.000	22:43	
Subject Sample	# 2	22:46	No Sample Provided
Breath Volume	0.000	22:46	
Air Blank	0.000	22:46	
Control Test	0.080	22:47	
Air Blank	0.000	22:47	
Diagnostics Check	OK	22:47	

Data Download Date/Time: 07/25/2017 11:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000961

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STONE,THOMAS,A
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: STONE,THOMAS
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 06/08/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:43

SUBJECT'S LAST NAME: BRANDA

DATE/TIME OF TEST: 06/25/2017 04:15:57

Diagnostics Check	OK	04:09
Air Blank	0.000	04:10
Control Test	0.080	04:10
Air Blank	0.000	04:11
Subject Sample # 1	0.172	04:11
Breath Volume	3.175	04:11
Air Blank	0.000	04:12
Air Blank	0.000	04:13
Subject Sample # 2	0.165	04:14
Breath Volume	2.941	04:14
Air Blank	0.000	04:15
Control Test	0.079	04:15
Air Blank	0.000	04:15
Diagnostics Check	OK	04:15

Data Download Date/Time: 07/25/2017 11:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000961

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HELMS, MATTHEW, L
BREATH TEST OPERATOR AGENCY: Wakulla County SO
ARRESTING OFFICER: HANKS, KYLE
ARRESTING OFFICER AGENCY: Wakulla County SO
LAST AGENCY INSPECTION: 06/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 08:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:53

SUBJECT'S LAST NAME: BARWICK

DATE/TIME OF TEST: 06/27/2017 09:22:00

Diagnostics Check	OK	09:16
Air Blank	0.000	09:16
Control Test	0.080	09:16
Air Blank	0.000	09:17
Subject Sample # 1	0.000	09:17
Breath Volume	2.382	09:17
Air Blank	0.000	09:18
Air Blank	0.000	09:20
Subject Sample # 2	0.000	09:20
Breath Volume	2.394	09:20
Air Blank	0.000	09:21
Control Test	0.079	09:21
Air Blank	0.000	09:21
Diagnostics Check	OK	09:22

Data Download Date/Time: 07/25/2017 11:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDEN,WILLIAM,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ALMEYDA,ERIC
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 18:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:30

SUBJECT'S LAST NAME: MORA

DATE/TIME OF TEST: 06/05/2017 19:55:51

Diagnostics Check	OK	19:51	
Air Blank	0.000	19:51	
Control Test	0.079	19:51	
Air Blank	0.000	19:52	
Subject Sample	# 1	19:54	Subject Test Refused
Breath Volume	0.000	19:54	
Air Blank	0.000	19:54	
Control Test	0.080	19:55	
Air Blank	0.000	19:55	
Diagnostics Check	OK	19:55	

Data Download Date/Time: 06/08/2017 14:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDEN,WILLIAM,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SPANN,JASON
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 05/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 18:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:27

SUBJECT'S LAST NAME: SALM

DATE/TIME OF TEST: 06/06/2017 19:55:48

Diagnostics Check	OK	19:49
Air Blank	0.000	19:49
Control Test	0.080	19:50
Air Blank	0.000	19:50
Subject Sample # 1	0.324	19:51
Breath Volume	1.691	19:51
Air Blank	0.000	19:51
Air Blank	0.000	19:53
Subject Sample # 2	0.327	19:54
Breath Volume	1.941	19:54
Air Blank	0.000	19:54
Control Test	0.079	19:55
Air Blank	0.000	19:55
Diagnostics Check	OK	19:55

Data Download Date/Time: 06/08/2017 14:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDEN,WILLIAM,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MCSPADDEN,JASON
ARRESTING OFFICER AGENCY: Winter Garden PD
LAST AGENCY INSPECTION: 05/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: LOVETT

DATE/TIME OF TEST: 06/08/2017 02:42:10

Diagnostics Check	OK	02:35
Air Blank	0.000	02:36
Control Test	0.081	02:36
Air Blank	0.000	02:36
Subject Sample # 1	0.198	02:37
Breath Volume	2.871	02:37
Air Blank	0.000	02:38
Air Blank	0.000	02:40
Subject Sample # 2	0.192	02:40
Breath Volume	3.054	02:40
Air Blank	0.000	02:41
Control Test	0.078	02:41
Air Blank	0.000	02:42
Diagnostics Check	OK	02:42

Data Download Date/Time: 06/08/2017 14:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDEN, WILLIAM, L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HALLOCK, STEPHEN
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 05/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: MCLAUGHLIN

DATE/TIME OF TEST: 06/09/2017 01:51:14

Diagnostics Check	OK	01:44
Air Blank	0.000	01:45
Control Test	0.080	01:45
Air Blank	0.000	01:46
Subject Sample # 1	0.214	01:46
Breath Volume	1.472	01:46
Air Blank	0.000	01:47
Air Blank	0.000	01:49
Subject Sample # 2	0.217	01:49
Breath Volume	1.859	01:49
Air Blank	0.000	01:50
Control Test	0.080	01:50
Air Blank	0.000	01:51
Diagnostics Check	OK	01:51

Data Download Date/Time: 06/14/2017 10:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDEN,WILLIAM,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: JORGE,JUNIOR
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:54

SUBJECT'S LAST NAME: DAHLMAN

DATE/TIME OF TEST: 06/13/2017 01:18:43

Diagnostics Check	OK	01:15	
Air Blank	0.000	01:15	
Control Test	0.081	01:16	
Air Blank	0.000	01:16	
Subject Sample	# 1	01:17	Subject Test Refused
Breath Volume	0.000	01:17	
Air Blank	0.000	01:17	
Control Test	0.080	01:18	
Air Blank	0.000	01:18	
Diagnostics Check	OK	01:18	

Data Download Date/Time: 06/14/2017 10:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDEN,WILLIAM,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: VINER,DENNIS
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 05/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 06/13/2017 03:58:50

Diagnostics Check	OK	03:50
Air Blank	0.000	03:50
Control Test	0.081	03:50
Air Blank	0.000	03:51
Subject Sample # 1	0.165	03:54
Breath Volume	2.078	03:54
Air Blank	0.000	03:54
Air Blank	0.000	03:56
Subject Sample # 2	0.162	03:57
Breath Volume	1.519	03:57
Air Blank	0.000	03:57
Control Test	0.078	03:58
Air Blank	0.000	03:58
Diagnostics Check	OK	03:58

Data Download Date/Time: 06/14/2017 10:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDEN, WILLIM, L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: EVANS, JOSH
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 16:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:05

SUBJECT'S LAST NAME: BURKE

DATE/TIME OF TEST: 06/13/2017 18:31:55

Diagnostics Check	OK	18:25
Air Blank	0.000	18:25
Control Test	0.080	18:26
Air Blank	0.000	18:26
Subject Sample # 1	0.193	18:27
Breath Volume	2.531	18:27
Air Blank	0.000	18:28
Air Blank	0.000	18:29
Subject Sample # 2	0.175	18:30
Breath Volume	2.738	18:30
Air Blank	0.000	18:31
Control Test	0.079	18:31
Air Blank	0.000	18:31
Diagnostics Check	OK	18:31

Data Download Date/Time: 06/14/2017 10:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDEN,WILLIAM,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: GONZALEZ,ELVIN
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 05/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: USSERY

DATE/TIME OF TEST: 06/14/2017 03:07:23

Diagnostics Check	OK	03:01
Air Blank	0.000	03:01
Control Test	0.080	03:01
Air Blank	0.000	03:02
Subject Sample # 1	0.234	03:02
Breath Volume	3.308	03:02
Air Blank	0.000	03:03
Air Blank	0.000	03:05
Subject Sample # 2	0.233	03:05
Breath Volume	2.246	03:05
Air Blank	0.000	03:06
Control Test	0.080	03:06
Air Blank	0.000	03:07
Diagnostics Check	OK	03:07

Data Download Date/Time: 06/14/2017 10:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,COURTNEY,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DRACH,STEFAN
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 06/14/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:33

SUBJECT'S LAST NAME: LARIN

DATE/TIME OF TEST: 06/15/2017 04:03:28

Diagnostics Check	OK	03:57
Air Blank	0.000	03:57
Control Test	0.080	03:58
Air Blank	0.000	03:58
Subject Sample # 1	0.179	03:59
Breath Volume	2.312	03:59
Air Blank	0.000	03:59
Air Blank	0.000	04:01
Subject Sample # 2	0.182	04:01
Breath Volume	2.589	04:01
Air Blank	0.000	04:02
Control Test	0.080	04:02
Air Blank	0.000	04:03
Diagnostics Check	OK	04:03

Data Download Date/Time: 06/22/2017 14:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: AZEVEDO,FABIO
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 06/14/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:52

SUBJECT'S LAST NAME: HENRY

DATE/TIME OF TEST: 06/16/2017 02:23:40

Diagnostics Check	OK	02:16
Air Blank	0.000	02:16
Control Test	0.081	02:16
Air Blank	0.000	02:17
Subject Sample # 1	0.012	02:19
Breath Volume	1.121	02:19
Air Blank	0.000	02:20
Air Blank	0.000	02:21
Subject Sample # 2	0.009	02:22
Breath Volume	1.898	02:22
Air Blank	0.000	02:22
Control Test	0.079	02:23
Air Blank	0.000	02:23
Diagnostics Check	OK	02:23

Data Download Date/Time: 06/22/2017 14:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: AZEVEDO,FABIO
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 06/14/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:54

SUBJECT'S LAST NAME: BONE

DATE/TIME OF TEST: 06/18/2017 03:26:53

Diagnostics Check	OK	03:20
Air Blank	0.000	03:21
Control Test	0.080	03:21
Air Blank	0.000	03:22
Subject Sample # 1	0.203	03:22
Breath Volume	2.355	03:22
Air Blank	0.000	03:23
Air Blank	0.000	03:24
Subject Sample # 2	0.208	03:25
Breath Volume	2.929	03:25
Air Blank	0.000	03:25
Control Test	0.079	03:26
Air Blank	0.000	03:26
Diagnostics Check	OK	03:26

Data Download Date/Time: 06/22/2017 14:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MOORE,ANDREW
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 06/14/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:58

SUBJECT'S LAST NAME: CATER

DATE/TIME OF TEST: 06/22/2017 05:27:56

Diagnostics Check	OK	05:20
Air Blank	0.000	05:21
Control Test	0.080	05:21
Air Blank	0.000	05:22
Subject Sample # 1	0.100	05:23
Breath Volume	1.820	05:23
Air Blank	0.000	05:24
Air Blank	0.000	05:26
Subject Sample # 2	0.097	05:26
Breath Volume	1.132	05:26
Air Blank	0.000	05:27
Control Test	0.080	05:27
Air Blank	0.000	05:27
Diagnostics Check	OK	05:27

Data Download Date/Time: 06/22/2017 14:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: RODRIGUEZ,JOSEPH
ARRESTING OFFICER AGENCY: Ocoee PD
LAST AGENCY INSPECTION: 06/14/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 07:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:40

SUBJECT'S LAST NAME: SWEETEN

DATE/TIME OF TEST: 06/23/2017 08:12:17

Diagnostics Check		OK	08:06
Air Blank		0.000	08:06
Control Test		0.081	08:07
Air Blank		0.000	08:07
Subject Sample	# 1	0.207	08:07
Breath Volume		2.648	08:07
Air Blank		0.000	08:08
Air Blank		0.000	08:10
Subject Sample	# 2	0.208	08:10
Breath Volume		1.613	08:10
Air Blank		0.000	08:11
Control Test		0.079	08:11
Air Blank		0.000	08:12
Diagnostics Check		OK	08:12

Data Download Date/Time: 06/29/2017 14:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ, MICHAEL, J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BORRAS, TAMARA
ARRESTING OFFICER AGENCY: Winter Garden PD
LAST AGENCY INSPECTION: 06/14/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 10:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:05

SUBJECT'S LAST NAME: WOHLGEMUTH

DATE/TIME OF TEST: 06/23/2017 11:38:49

Diagnostics Check	OK	11:32
Air Blank	0.000	11:32
Control Test	0.080	11:33
Air Blank	0.000	11:33
Subject Sample # 1	0.000	11:34
Breath Volume	1.531	11:34
Air Blank	0.000	11:34
Air Blank	0.000	11:36
Subject Sample # 2	0.000	11:37
Breath Volume	2.054	11:37
Air Blank	0.000	11:37
Control Test	0.080	11:38
Air Blank	0.000	11:38
Diagnostics Check	OK	11:38

Data Download Date/Time: 06/29/2017 14:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,CALEB,N
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: AZEVEDO,FABIO
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 06/14/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 06/24/2017 03:41:50

Diagnostics Check		OK	03:29	
Air Blank		0.000	03:29	
Control Test		0.081	03:29	
Air Blank		0.000	03:30	
Subject Sample	# 1	0.093	03:33	Volume Not Met
Breath Volume		0.328	03:33	
Air Blank		0.000	03:34	
Air Blank		0.000	03:36	
Subject Sample	# 2	0.094	03:37	
Breath Volume		1.121	03:37	
Air Blank		0.000	03:37	
Air Blank		0.000	03:39	
Subject Sample	# 3	0.098	03:40	
Breath Volume		1.117	03:40	
Air Blank		0.000	03:40	
Control Test		0.079	03:41	
Air Blank		0.000	03:41	
Diagnostics Check		OK	03:41	

Data Download Date/Time: 06/29/2017 14:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: THOMPSON,JOSEPH
ARRESTING OFFICER AGENCY: Winter Garden PD
LAST AGENCY INSPECTION: 06/14/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:16

SUBJECT'S LAST NAME: PERSAUD

DATE/TIME OF TEST: 06/25/2017 04:51:08

Diagnostics Check		OK	04:42
Air Blank		0.000	04:42
Control Test		0.081	04:43
Air Blank		0.000	04:43
Subject Sample	# 1	0.309	04:45
Breath Volume		1.679	04:45
Air Blank		0.000	04:45
Air Blank		0.000	04:47
Subject Sample	# 2	0.291	04:49
Breath Volume		1.687	04:49
Air Blank		0.000	04:50
Control Test		0.078	04:50
Air Blank		0.000	04:51
Diagnostics Check		OK	04:51

Data Download Date/Time: 06/29/2017 14:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ROY,BRYAN
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 06/14/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 08:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:55

SUBJECT'S LAST NAME: SAEZ

DATE/TIME OF TEST: 06/25/2017 09:25:45

Diagnostics Check	OK	09:19
Air Blank	0.000	09:20
Control Test	0.079	09:20
Air Blank	0.000	09:21
Subject Sample # 1	0.000	09:21
Breath Volume	2.601	09:21
Air Blank	0.000	09:21
Air Blank	0.000	09:23
Subject Sample # 2	0.000	09:24
Breath Volume	3.371	09:24
Air Blank	0.000	09:24
Control Test	0.080	09:25
Air Blank	0.000	09:25
Diagnostics Check	OK	09:25

Data Download Date/Time: 06/29/2017 14:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000965

INSTRUMENT REGISTERED TO: QUINCY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FINUFF,DANIEL,W
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: FINUFF,DANIEL
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 05/09/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:26

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 06/03/2017 01:53:08

Diagnostics Check	OK	01:47
Air Blank	0.000	01:47
Control Test	0.079	01:47
Air Blank	0.000	01:48
Subject Sample # 1	0.000	01:48
Breath Volume	1.855	01:48
Air Blank	0.000	01:49
Air Blank	0.000	01:51
Subject Sample # 2	0.000	01:51
Breath Volume	1.859	01:51
Air Blank	0.000	01:52
Control Test	0.079	01:52
Air Blank	0.000	01:53
Diagnostics Check	OK	01:53

Data Download Date/Time: 06/23/2017 08:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000965

INSTRUMENT REGISTERED TO: QUINCY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLEN,JAMES,R
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: ALLEN,JAMES
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 05/09/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 06/04/2017 02:02:14

Diagnostics Check	OK	01:55
Air Blank	0.000	01:56
Control Test	0.077	01:56
Air Blank	0.000	01:57
Subject Sample # 1	0.158	01:57
Breath Volume	1.992	01:57
Air Blank	0.000	01:58
Air Blank	0.000	02:00
Subject Sample # 2	0.163	02:00
Breath Volume	1.992	02:00
Air Blank	0.000	02:01
Control Test	0.078	02:01
Air Blank	0.000	02:02
Diagnostics Check	OK	02:02

Data Download Date/Time: 06/23/2017 08:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000965

INSTRUMENT REGISTERED TO: QUINCY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLEN,JAMES,R
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: ALLEN,JAMES
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 05/09/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 06/13/2017 03:02:47

Diagnostics Check	OK	02:56
Air Blank	0.000	02:56
Control Test	0.078	02:57
Air Blank	0.000	02:57
Subject Sample # 1	0.120	02:58
Breath Volume	2.859	02:58
Air Blank	0.000	02:58
Air Blank	0.000	03:00
Subject Sample # 2	0.117	03:01
Breath Volume	2.683	03:01
Air Blank	0.000	03:01
Control Test	0.078	03:02
Air Blank	0.000	03:02
Diagnostics Check	OK	03:02

Data Download Date/Time: 06/23/2017 08:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000965

INSTRUMENT REGISTERED TO: QUINCY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COPELAND,MICHAEL,C
BREATH TEST OPERATOR AGENCY: Quincy PD
ARRESTING OFFICER: COPELAND,MICHAEL
ARRESTING OFFICER AGENCY: Quincy PD
LAST AGENCY INSPECTION: 05/09/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:01

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 06/16/2017 03:29:12

Diagnostics Check	OK	03:22
Air Blank	0.000	03:22
Control Test	0.078	03:23
Air Blank	0.000	03:23
Subject Sample # 1	0.180	03:24
Breath Volume	1.371	03:24
Air Blank	0.000	03:25
Air Blank	0.000	03:27
Subject Sample # 2	0.182	03:27
Breath Volume	1.417	03:27
Air Blank	0.000	03:28
Control Test	0.078	03:28
Air Blank	0.000	03:29
Diagnostics Check	OK	03:29

Data Download Date/Time: 06/23/2017 08:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000965

INSTRUMENT REGISTERED TO: QUINCY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARROW,JESHUA,C
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: BARROW,JESHUA
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 05/09/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:37

SUBJECT'S LAST NAME: CALVIN

DATE/TIME OF TEST: 06/16/2017 22:04:00

Diagnostics Check	OK	21:57
Air Blank	0.000	21:58
Control Test	0.077	21:58
Air Blank	0.000	21:59
Subject Sample # 1	0.000	21:59
Breath Volume	1.316	21:59
Air Blank	0.000	22:00
Air Blank	0.000	22:02
Subject Sample # 2	0.000	22:02
Breath Volume	1.578	22:02
Air Blank	0.000	22:03
Control Test	0.077	22:03
Air Blank	0.000	22:03
Diagnostics Check	OK	22:04

Data Download Date/Time: 06/23/2017 08:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000965

INSTRUMENT REGISTERED TO: QUINCY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLEN,JAMES,R
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: ALLEN,JAMES
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 05/09/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: CONYERS

DATE/TIME OF TEST: 06/22/2017 02:18:20

Diagnostics Check	OK	02:10
Air Blank	0.000	02:11
Control Test	0.078	02:11
Air Blank	0.000	02:12
Subject Sample # 1	0.173	02:13
Breath Volume	2.093	02:13
Air Blank	0.000	02:14
Air Blank	0.000	02:15
Subject Sample # 2	0.165	02:16
Breath Volume	1.187	02:16
Air Blank	0.000	02:17
Control Test	0.077	02:17
Air Blank	0.000	02:18
Diagnostics Check	OK	02:18

Data Download Date/Time: 06/23/2017 08:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: APOPKA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPBELL,ROBERT,J
BREATH TEST OPERATOR AGENCY: Apopka PD
ARRESTING OFFICER: CAMPBELL,ROBERT
ARRESTING OFFICER AGENCY: Apopka PD
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:23

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 06/05/2017 03:06:50

Diagnostics Check	OK	02:51	
Air Blank	0.000	02:51	
Control Test	0.079	02:51	
Air Blank	0.000	02:52	
Subject Sample	# 1 0.159	02:54	
Breath Volume	1.222	02:54	
Air Blank	0.000	02:54	
Air Blank	0.000	02:56	
Subject Sample	# 2	02:59	No Sample Provided
Breath Volume	0.000	02:59	
Air Blank	0.000	03:00	
Air Blank	0.000	03:02	
Subject Sample	# 3	03:05	No Sample Provided
Breath Volume	0.000	03:05	
Air Blank	0.000	03:05	
Control Test	0.077	03:06	
Air Blank	0.000	03:06	
Diagnostics Check	OK	03:06	

Data Download Date/Time: 08/08/2017 10:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: APOPKA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPBELL,ROBERT,J
BREATH TEST OPERATOR AGENCY: Apopka PD
ARRESTING OFFICER: CAMPBELL,ROBERT
ARRESTING OFFICER AGENCY: Apopka PD
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:21

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 06/05/2017 03:59:07

Diagnostics Check	OK	03:47	
Air Blank	0.000	03:48	
Control Test	0.079	03:48	
Air Blank	0.000	03:49	
Subject Sample	# 1	03:52	No Sample Provided
Breath Volume	0.000	03:52	
Air Blank	0.000	03:52	
Air Blank	0.000	03:54	
Subject Sample	# 2	03:57	No Sample Provided
Breath Volume	0.000	03:57	
Air Blank	0.000	03:58	
Control Test	0.078	03:58	
Air Blank	0.000	03:59	
Diagnostics Check	OK	03:59	

Data Download Date/Time: 08/08/2017 10:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: APOPKA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLER,ASHLEY,B
BREATH TEST OPERATOR AGENCY: Apopka PD
ARRESTING OFFICER: FEQUIRE,MACINTOSH
ARRESTING OFFICER AGENCY: Apopka PD
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 11:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 12:28

SUBJECT'S LAST NAME: ROBINSON

DATE/TIME OF TEST: 06/11/2017 13:04:02

Diagnostics Check	OK	12:52	
Air Blank	0.000	12:53	
Control Test	0.078	12:53	
Air Blank	0.000	12:53	
Subject Sample	# 1	12:57	No Sample Provided
Breath Volume	0.000	12:57	
Air Blank	0.000	12:57	
Air Blank	0.000	12:59	
Subject Sample	# 2	13:02	No Sample Provided
Breath Volume	0.000	13:02	
Air Blank	0.000	13:03	
Control Test	0.078	13:03	
Air Blank	0.000	13:03	
Diagnostics Check	OK	13:04	

Data Download Date/Time: 08/08/2017 10:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: APOPKA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLER,ASHLEY,B
BREATH TEST OPERATOR AGENCY: Apopka PD
ARRESTING OFFICER: FEQUIRE,MACINTOSH
ARRESTING OFFICER AGENCY: Apopka PD
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 11:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 12:43

SUBJECT'S LAST NAME: ROBINSON

DATE/TIME OF TEST: 06/11/2017 13:24:01

Diagnostics Check	OK	13:16
Air Blank	0.000	13:16
Control Test	0.078	13:17
Air Blank	0.000	13:17
Subject Sample # 1	0.000	13:19
Breath Volume	3.308	13:19
Air Blank	0.000	13:20
Air Blank	0.000	13:22
Subject Sample # 2	0.000	13:22
Breath Volume	3.164	13:22
Air Blank	0.000	13:23
Control Test	0.079	13:23
Air Blank	0.000	13:23
Diagnostics Check	OK	13:24

Data Download Date/Time: 08/08/2017 10:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: APOPKA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPBELL,ROBERT,J
BREATH TEST OPERATOR AGENCY: Apopka PD
ARRESTING OFFICER: CAMPBELL,ROBERT
ARRESTING OFFICER AGENCY: Apopka PD
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: CAJUSTE

DATE/TIME OF TEST: 06/14/2017 01:05:12

Diagnostics Check	OK	00:59
Air Blank	0.000	00:59
Control Test	0.079	01:00
Air Blank	0.000	01:00
Subject Sample # 1	0.000	01:01
Breath Volume	3.789	01:01
Air Blank	0.000	01:01
Air Blank	0.000	01:03
Subject Sample # 2	0.000	01:03
Breath Volume	3.171	01:03
Air Blank	0.000	01:04
Control Test	0.080	01:04
Air Blank	0.000	01:05
Diagnostics Check	OK	01:05

Data Download Date/Time: 08/08/2017 10:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: APOPKA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPBELL,ROBERT,J
BREATH TEST OPERATOR AGENCY: Apopka PD
ARRESTING OFFICER: CAMPBELL,ROBERT
ARRESTING OFFICER AGENCY: Apopka PD
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: KELLY

DATE/TIME OF TEST: 06/15/2017 03:40:50

Diagnostics Check	OK	03:34
Air Blank	0.000	03:35
Control Test	0.078	03:35
Air Blank	0.000	03:36
Subject Sample # 1	0.029	03:36
Breath Volume	1.296	03:36
Air Blank	0.000	03:37
Air Blank	0.000	03:38
Subject Sample # 2	0.028	03:39
Breath Volume	1.152	03:39
Air Blank	0.000	03:39
Control Test	0.078	03:40
Air Blank	0.000	03:40
Diagnostics Check	OK	03:40

Data Download Date/Time: 08/08/2017 10:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: APOPKA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLER,ASHLEY,B
BREATH TEST OPERATOR AGENCY: Apopka PD
ARRESTING OFFICER: ELLER,ASHLEY
ARRESTING OFFICER AGENCY: Apopka PD
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:46

SUBJECT'S LAST NAME: RODGERS

DATE/TIME OF TEST: 06/26/2017 00:15:41

Diagnostics Check		OK	00:07
Air Blank		0.000	00:07
Control Test		0.078	00:07
Air Blank		0.000	00:08
Subject Sample	# 1	0.170	00:10
Breath Volume		1.742	00:10
Air Blank		0.000	00:11
Air Blank		0.000	00:12
Subject Sample	# 2	0.183	00:14
Breath Volume		1.695	00:14
Air Blank		0.000	00:14
Control Test		0.078	00:15
Air Blank		0.000	00:15
Diagnostics Check		OK	00:15

Data Download Date/Time: 08/08/2017 10:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: APOPKA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLER,ASHLEY,B
BREATH TEST OPERATOR AGENCY: Apopka PD
ARRESTING OFFICER: ELLER,ASHLEY
ARRESTING OFFICER AGENCY: Apopka PD
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:23

SUBJECT'S LAST NAME: FREEMAN

DATE/TIME OF TEST: 06/27/2017 03:57:14

Diagnostics Check		OK	03:50
Air Blank		0.000	03:51
Control Test		0.079	03:51
Air Blank		0.000	03:51
Subject Sample	# 1	0.154	03:52
Breath Volume		4.671	03:52
Air Blank		0.000	03:53
Air Blank		0.000	03:55
Subject Sample	# 2	0.153	03:55
Breath Volume		4.796	03:55
Air Blank		0.000	03:56
Control Test		0.078	03:56
Air Blank		0.000	03:57
Diagnostics Check		OK	03:57

Data Download Date/Time: 08/08/2017 10:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: APOPKA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIARRATANA,JOHN,R
BREATH TEST OPERATOR AGENCY: Apopka PD
ARRESTING OFFICER: MELLON,JENNIFER
ARRESTING OFFICER AGENCY: Apopka PD
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:54

SUBJECT'S LAST NAME: LAMONS

DATE/TIME OF TEST: 06/28/2017 04:29:30

Diagnostics Check	OK	04:23
Air Blank	0.000	04:23
Control Test	0.078	04:24
Air Blank	0.000	04:24
Subject Sample # 1	0.180	04:25
Breath Volume	3.367	04:25
Air Blank	0.000	04:25
Air Blank	0.000	04:27
Subject Sample # 2	0.170	04:27
Breath Volume	3.343	04:27
Air Blank	0.000	04:28
Control Test	0.078	04:28
Air Blank	0.000	04:29
Diagnostics Check	OK	04:29

Data Download Date/Time: 08/08/2017 10:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPOS, MICHAEL, A
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: MARTINEZ, YOMARI
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 05/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 02:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:01

SUBJECT'S LAST NAME: SANABRIA

DATE/TIME OF TEST: 06/03/2017 03:51:01

Diagnostics Check	OK	03:44
Air Blank	0.000	03:45
Control Test	0.080	03:45
Air Blank	0.000	03:46
Subject Sample # 1	0.247	03:46
Breath Volume	1.671	03:46
Air Blank	0.000	03:47
Air Blank	0.000	03:48
Subject Sample # 2	0.265	03:49
Breath Volume	1.582	03:49
Air Blank	0.000	03:50
Control Test	0.079	03:50
Air Blank	0.000	03:50
Diagnostics Check	OK	03:51

Data Download Date/Time: 06/27/2017 20:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPOS,MICHAEL,A
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: CAMPOS,MICHAEL
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 05/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 03:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: RIGATUSO

DATE/TIME OF TEST: 06/03/2017 04:01:57

Diagnostics Check	OK	03:55
Air Blank	0.000	03:56
Control Test	0.080	03:56
Air Blank	0.000	03:56
Subject Sample # 1	0.256	03:57
Breath Volume	1.824	03:57
Air Blank	0.000	03:58
Air Blank	0.000	04:00
Subject Sample # 2	0.256	04:00
Breath Volume	1.898	04:00
Air Blank	0.000	04:01
Control Test	0.078	04:01
Air Blank	0.000	04:01
Diagnostics Check	OK	04:01

Data Download Date/Time: 06/27/2017 20:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPOS, MICHAEL, A
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: CAMPOS, MICHAEL
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 05/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:21

SUBJECT'S LAST NAME: HANLEY

DATE/TIME OF TEST: 06/04/2017 02:47:28

Diagnostics Check	OK	02:41
Air Blank	0.000	02:41
Control Test	0.081	02:42
Air Blank	0.000	02:42
Subject Sample # 1	0.180	02:43
Breath Volume	4.496	02:43
Air Blank	0.000	02:43
Air Blank	0.000	02:45
Subject Sample # 2	0.183	02:45
Breath Volume	4.726	02:45
Air Blank	0.000	02:46
Control Test	0.080	02:46
Air Blank	0.000	02:47
Diagnostics Check	OK	02:47

Data Download Date/Time: 06/27/2017 20:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEADOWS,CHARLES,W
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: MCCONNELL,CHARLES
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 05/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 13:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:06

SUBJECT'S LAST NAME: MEADOWS

DATE/TIME OF TEST: 06/07/2017 14:34:27

Diagnostics Check	OK	14:28
Air Blank	0.000	14:28
Control Test	0.080	14:28
Air Blank	0.000	14:29
Subject Sample # 1	0.000	14:29
Breath Volume	3.046	14:29
Air Blank	0.000	14:30
Air Blank	0.000	14:32
Subject Sample # 2	0.000	14:33
Breath Volume	3.148	14:33
Air Blank	0.000	14:33
Control Test	0.080	14:33
Air Blank	0.000	14:34
Diagnostics Check	OK	14:34

Data Download Date/Time: 06/27/2017 20:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STINSON,ANDREW,L
BREATH TEST OPERATOR AGENCY: Kissimmee PD
ARRESTING OFFICER: STINSON,ANDREW
ARRESTING OFFICER AGENCY: Kissimmee PD
LAST AGENCY INSPECTION: 05/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 01:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:11

SUBJECT'S LAST NAME: CARBONELL

DATE/TIME OF TEST: 06/12/2017 01:38:39

Diagnostics Check	OK	01:32
Air Blank	0.000	01:32
Control Test	0.080	01:32
Air Blank	0.000	01:33
Subject Sample # 1	0.164	01:34
Breath Volume	1.445	01:34
Air Blank	0.000	01:34
Air Blank	0.000	01:36
Subject Sample # 2	0.150	01:37
Breath Volume	2.886	01:37
Air Blank	0.000	01:37
Control Test	0.079	01:38
Air Blank	0.000	01:38
Diagnostics Check	OK	01:38

Data Download Date/Time: 06/27/2017 20:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TELLERIA,TROY,R
BREATH TEST OPERATOR AGENCY: St. Cloud PD
ARRESTING OFFICER: CARLO,NATHANIAL
ARRESTING OFFICER AGENCY: St. Cloud PD
LAST AGENCY INSPECTION: 05/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 00:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: ROGERS

DATE/TIME OF TEST: 06/14/2017 00:56:36

Diagnostics Check		OK	00:41	
Air Blank		0.000	00:41	
Control Test		0.081	00:41	
Air Blank		0.000	00:42	
Subject Sample	# 1	0.197	00:45	Volume Not Met
Breath Volume		0.261	00:45	
Air Blank		0.000	00:46	
Air Blank		0.000	00:48	
Subject Sample	# 2	0.196	00:49	
Breath Volume		1.648	00:49	
Air Blank		0.000	00:50	
Air Blank		0.000	00:51	
Subject Sample	# 3	0.167	00:55	Volume Not Met
Breath Volume		0.300	00:55	
Air Blank		0.000	00:55	
Control Test		0.078	00:55	
Air Blank		0.000	00:56	
Diagnostics Check		OK	00:56	

Data Download Date/Time: 06/27/2017 20:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPOS,MICHAEL,A
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: CAMPOS,MICHAEL
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 05/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 01:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:47

SUBJECT'S LAST NAME: WINSTON

DATE/TIME OF TEST: 06/16/2017 02:31:19

Diagnostics Check	OK	02:19	
Air Blank	0.000	02:19	
Control Test	0.080	02:20	
Air Blank	0.000	02:20	
Subject Sample	# 1 0.193	02:21	
Breath Volume	1.992	02:21	
Air Blank	0.000	02:21	
Air Blank	0.000	02:23	
Subject Sample	# 2	02:26	No Sample Provided
Breath Volume	0.000	02:26	
Air Blank	0.000	02:27	
Air Blank	0.000	02:29	
Subject Sample	# 3 0.185	02:29	
Breath Volume	2.105	02:29	
Air Blank	0.000	02:30	
Control Test	0.080	02:30	
Air Blank	0.000	02:31	
Diagnostics Check	OK	02:31	

Data Download Date/Time: 06/27/2017 20:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUSSAK,KEVIN,M
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: RUSSAK,KEVIN
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 04:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:25

SUBJECT'S LAST NAME: EVERETT

DATE/TIME OF TEST: 06/17/2017 05:52:55

Diagnostics Check	OK	05:45	
Air Blank	0.000	05:46	
Control Test	0.081	05:46	
Air Blank	0.000	05:47	
Subject Sample	# 1 0.122	05:48	
Breath Volume	1.414	05:48	
Air Blank	0.000	05:48	
Air Blank	0.000	05:50	
Subject Sample	# 2 0.138	05:51	
Breath Volume	2.132	05:51	
Air Blank	0.000	05:51	
Control Test	0.080	05:52	
Air Blank		05:52	RFI Detect
Air Blank	0.000	05:52	

Data Download Date/Time: 06/27/2017 20:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUSSAK,KEVIN,M
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: RUSSAK,KEVIN
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 04:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:25

SUBJECT'S LAST NAME: EVERETT

DATE/TIME OF TEST: 06/17/2017 06:02:24

Diagnostics Check	OK	05:55
Air Blank	0.000	05:55
Control Test	0.080	05:55
Air Blank	0.000	05:56
Subject Sample # 1	0.135	05:57
Breath Volume	1.683	05:57
Air Blank	0.000	05:57
Air Blank	0.000	05:59
Subject Sample # 2	0.125	06:00
Breath Volume	2.351	06:00
Air Blank	0.000	06:01
Control Test	0.080	06:01
Air Blank	0.000	06:02
Diagnostics Check	OK	06:02

Data Download Date/Time: 06/27/2017 20:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCONNELL,CHARLES,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: MCCONNELL,CHARLES
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 05/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 16:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:20

SUBJECT'S LAST NAME: CALDERON-TORRES

DATE/TIME OF TEST: 06/18/2017 17:52:09

Diagnostics Check	OK	17:45
Air Blank	0.000	17:46
Control Test	0.080	17:46
Air Blank	0.000	17:47
Subject Sample # 1	0.064	17:47
Breath Volume	2.328	17:47
Air Blank	0.000	17:48
Air Blank	0.000	17:50
Subject Sample # 2	0.066	17:50
Breath Volume	2.656	17:50
Air Blank	0.000	17:51
Control Test	0.080	17:51
Air Blank	0.000	17:52
Diagnostics Check	OK	17:52

Data Download Date/Time: 06/27/2017 20:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LYONS,D,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: LYONS,D
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 05/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 21:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:08

SUBJECT'S LAST NAME: ESPINAL

DATE/TIME OF TEST: 06/20/2017 22:34:18

Diagnostics Check	OK	22:30	
Air Blank	0.000	22:30	
Control Test	0.080	22:31	
Air Blank	0.000	22:31	
Subject Sample	# 1 0.208	22:32	
Breath Volume	2.808	22:32	
Air Blank		22:32	RFI Detect
Air Blank	0.000	22:33	
Control Test	0.080	22:33	
Air Blank	0.000	22:34	
Diagnostics Check	OK	22:34	

Data Download Date/Time: 06/27/2017 20:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LYONS,D,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: LYONS,D
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 05/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 21:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:08

SUBJECT'S LAST NAME: ESPINAL

DATE/TIME OF TEST: 06/20/2017 22:43:32

Diagnostics Check	OK	22:36
Air Blank	0.000	22:37
Control Test	0.080	22:37
Air Blank	0.000	22:38
Subject Sample # 1	0.198	22:39
Breath Volume	1.972	22:39
Air Blank	0.000	22:39
Air Blank	0.000	22:41
Subject Sample # 2	0.198	22:42
Breath Volume	2.171	22:42
Air Blank	0.000	22:42
Control Test	0.077	22:42
Air Blank	0.000	22:43
Diagnostics Check	OK	22:43

Data Download Date/Time: 06/27/2017 20:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESH,JOSH,A
BREATH TEST OPERATOR AGENCY: St. Cloud PD
ARRESTING OFFICER: ESH,JOSH
ARRESTING OFFICER AGENCY: St. Cloud PD
LAST AGENCY INSPECTION: 05/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 22:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: RAMANAND

DATE/TIME OF TEST: 06/21/2017 22:56:53

Diagnostics Check	OK	22:50
Air Blank	0.000	22:51
Control Test	0.080	22:51
Air Blank	0.000	22:52
Subject Sample # 1	0.227	22:52
Breath Volume	1.484	22:52
Air Blank	0.000	22:53
Air Blank	0.000	22:54
Subject Sample # 2	0.221	22:55
Breath Volume	2.687	22:55
Air Blank	0.000	22:55
Control Test	0.079	22:56
Air Blank	0.000	22:56
Diagnostics Check	OK	22:56

Data Download Date/Time: 06/27/2017 20:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESH,JOSH,A
BREATH TEST OPERATOR AGENCY: St. Cloud PD
ARRESTING OFFICER: ESH,JOSH
ARRESTING OFFICER AGENCY: St. Cloud PD
LAST AGENCY INSPECTION: 05/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: ZERO TOL
ARREST TIME: 00:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:36

SUBJECT'S LAST NAME: COOLEY

DATE/TIME OF TEST: 06/24/2017 01:16:45

Diagnostics Check	OK	01:09
Air Blank	0.000	01:10
Control Test	0.080	01:10
Air Blank	0.000	01:10
Subject Sample # 1	0.238	01:11
Breath Volume	1.546	01:11
Air Blank	0.000	01:12
Air Blank	0.000	01:14
Subject Sample # 2	0.230	01:15
Breath Volume	1.121	01:15
Air Blank	0.000	01:15
Control Test	0.080	01:16
Air Blank	0.000	01:16
Diagnostics Check	OK	01:16

Data Download Date/Time: 06/27/2017 20:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPOS,MICHAEL,A
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: CAMPOS,MICHAEL
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 05/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 02:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 06/24/2017 03:28:57

Diagnostics Check	OK	03:21
Air Blank	0.000	03:22
Control Test	0.080	03:22
Air Blank	0.000	03:22
Subject Sample # 1	0.118	03:24
Breath Volume	2.007	03:24
Air Blank	0.000	03:25
Air Blank	0.000	03:26
Subject Sample # 2	0.122	03:27
Breath Volume	2.015	03:27
Air Blank	0.000	03:28
Control Test	0.078	03:28
Air Blank	0.000	03:28
Diagnostics Check	OK	03:28

Data Download Date/Time: 06/27/2017 20:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GAYLE,OWEN,D
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: MINERVINO,KRISTIN
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 05/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 07:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:17

SUBJECT'S LAST NAME: GALINDO

DATE/TIME OF TEST: 06/24/2017 08:44:02

Diagnostics Check		OK	08:37
Air Blank		0.000	08:37
Control Test		0.079	08:38
Air Blank		0.000	08:38
Subject Sample	# 1	0.214	08:38
Breath Volume		1.132	08:38
Air Blank		0.000	08:39
Air Blank		0.000	08:41
Subject Sample	# 2	0.231	08:42
Breath Volume		2.031	08:42
Air Blank		0.000	08:43
Control Test		0.079	08:43
Air Blank		0.000	08:43
Diagnostics Check		OK	08:44

Data Download Date/Time: 06/27/2017 20:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCONNELL,CHARLES,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: MCCONNELL,CHARLES
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 05/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 17:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:49

SUBJECT'S LAST NAME: MALAVE

DATE/TIME OF TEST: 06/27/2017 18:13:13

Diagnostics Check	OK	18:11	
Air Blank	0.000	18:11	
Control Test	0.080	18:11	
Air Blank		18:12	Purge Fail
Air Blank	0.000	18:13	

Data Download Date/Time: 06/27/2017 20:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BEEKMAN,JACOB,C
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: MCCONNELL,CHARLES
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 05/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 17:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:49

SUBJECT'S LAST NAME: MALAVE

DATE/TIME OF TEST: 06/27/2017 18:25:11

Diagnostics Check	OK	18:18
Air Blank	0.000	18:18
Control Test	0.080	18:19
Air Blank	0.000	18:19
Subject Sample # 1	0.000	18:20
Breath Volume	2.015	18:20
Air Blank	0.000	18:21
Air Blank	0.000	18:23
Subject Sample # 2	0.000	18:23
Breath Volume	3.015	18:23
Air Blank	0.000	18:24
Control Test	0.081	18:24
Air Blank	0.000	18:25
Diagnostics Check	OK	18:25

Data Download Date/Time: 06/27/2017 20:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETTS, RONALD, E
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: HOBUS, CAVIN
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 05/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 14:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:55

SUBJECT'S LAST NAME: MAKI

DATE/TIME OF TEST: 06/13/2017 15:25:27

Diagnostics Check	OK	15:19
Air Blank	0.000	15:19
Control Test	0.077	15:19
Air Blank	0.000	15:20
Subject Sample # 1	0.142	15:21
Breath Volume	2.089	15:21
Air Blank	0.000	15:21
Air Blank	0.000	15:23
Subject Sample # 2	0.137	15:24
Breath Volume	1.632	15:24
Air Blank	0.000	15:24
Control Test	0.078	15:24
Air Blank	0.000	15:25
Diagnostics Check	OK	15:25

Data Download Date/Time: 06/29/2017 10:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHEINER,JAMES,R
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: SCHEINER,JAMES
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 05/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 36314080A3
DRY GAS EXPIRY DATE: 02/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: HOXIE

DATE/TIME OF TEST: 06/19/2017 00:30:20

Diagnostics Check	OK	00:24
Air Blank	0.000	00:25
Control Test	0.079	00:25
Air Blank	0.000	00:25
Subject Sample # 1	0.129	00:26
Breath Volume	2.750	00:26
Air Blank	0.000	00:26
Air Blank	0.000	00:28
Subject Sample # 2	0.135	00:28
Breath Volume	1.871	00:28
Air Blank	0.000	00:29
Control Test	0.079	00:29
Air Blank	0.000	00:30
Diagnostics Check	OK	00:30

Data Download Date/Time: 06/29/2017 10:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUINN,DAVID,J
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: MEEHAN,RYAN
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 05/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 36314080A3
DRY GAS EXPIRY DATE: 02/05/2017
VIOLATION CODE: DUI
ARREST TIME: 17:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:50

SUBJECT'S LAST NAME: BECKFORD

DATE/TIME OF TEST: 06/23/2017 08:05:36

Diagnostics Check	OK	07:59
Air Blank	0.000	07:59
Control Test	0.079	08:00
Air Blank	0.000	08:00
Subject Sample # 1	0.191	08:01
Breath Volume	1.527	08:01
Air Blank	0.000	08:01
Air Blank	0.000	08:03
Subject Sample # 2	0.185	08:04
Breath Volume	2.292	08:04
Air Blank	0.000	08:04
Control Test	0.079	08:05
Air Blank	0.000	08:05
Diagnostics Check	OK	08:05

Data Download Date/Time: 06/29/2017 10:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLFER,KEVIN,A
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: VOLK,N
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 06/23/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 36314080A3
DRY GAS EXPIRY DATE: 02/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: COLFER

DATE/TIME OF TEST: 06/23/2017 10:49:35

Diagnostics Check	OK	10:42
Air Blank	0.000	10:43
Control Test	0.079	10:43
Air Blank	0.000	10:44
Subject Sample # 1	0.205	10:45
Breath Volume	5.492	10:45
Air Blank	0.000	10:45
Air Blank	0.000	10:47
Subject Sample # 2	0.197	10:48
Breath Volume	4.945	10:48
Air Blank	0.000	10:48
Control Test	0.079	10:49
Air Blank	0.000	10:49
Diagnostics Check	OK	10:49

Data Download Date/Time: 08/02/2017 13:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHEINER,JAMES,R
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: SCHEINER,JAMES
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 06/23/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 36314080A3
DRY GAS EXPIRY DATE: 02/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:02

SUBJECT'S LAST NAME: COLVIN

DATE/TIME OF TEST: 06/23/2017 11:36:28

Diagnostics Check	OK	11:29
Air Blank	0.000	11:30
Control Test	0.080	11:30
Air Blank	0.000	11:30
Subject Sample # 1	0.160	11:31
Breath Volume	1.484	11:31
Air Blank	0.000	11:32
Air Blank	0.000	11:33
Subject Sample # 2	0.163	11:35
Breath Volume	1.726	11:35
Air Blank	0.000	11:35
Control Test	0.080	11:35
Air Blank	0.000	11:36
Diagnostics Check	OK	11:36

Data Download Date/Time: 08/02/2017 13:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

INSTRUMENT REGISTERED TO: TITUSVILLE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCGRORY,DIARMUID,J
BREATH TEST OPERATOR AGENCY: Titusville PD
ARRESTING OFFICER: MCGRORY,DIARMUID
ARRESTING OFFICER AGENCY: Titusville PD
LAST AGENCY INSPECTION: 05/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:46

SUBJECT'S LAST NAME: ELROD

DATE/TIME OF TEST: 06/15/2017 22:35:43

Diagnostics Check	OK	22:28
Air Blank	0.000	22:29
Control Test	0.078	22:29
Air Blank	0.000	22:30
Subject Sample # 1	0.162	22:31
Breath Volume	3.136	22:31
Air Blank	0.000	22:31
Air Blank	0.000	22:33
Subject Sample # 2	0.161	22:34
Breath Volume	2.046	22:34
Air Blank	0.000	22:34
Control Test	0.078	22:35
Air Blank	0.000	22:35
Diagnostics Check	OK	22:35

Data Download Date/Time: 06/16/2017 21:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

INSTRUMENT REGISTERED TO: TITUSVILLE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLDEN,ROBERT,A
BREATH TEST OPERATOR AGENCY: Titusville PD
ARRESTING OFFICER: HOLDEN,ROBERT
ARRESTING OFFICER AGENCY: Titusville PD
LAST AGENCY INSPECTION: 06/16/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 17:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:44

SUBJECT'S LAST NAME: MERZ

DATE/TIME OF TEST: 06/23/2017 17:51:17

Diagnostics Check	OK	17:44
Air Blank	0.000	17:44
Control Test	0.079	17:45
Air Blank	0.000	17:45
Subject Sample # 1	0.326	17:46
Breath Volume	1.738	17:46
Air Blank	0.000	17:47
Air Blank	0.000	17:48
Subject Sample # 2	0.321	17:49
Breath Volume	1.597	17:49
Air Blank	0.000	17:50
Control Test	0.079	17:50
Air Blank	0.000	17:51
Diagnostics Check	OK	17:51

Data Download Date/Time: 07/16/2017 12:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

INSTRUMENT REGISTERED TO: TITUSVILLE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCGRORY,DIARMUID,J
BREATH TEST OPERATOR AGENCY: Titusville PD
ARRESTING OFFICER: DURDEN,SHAWNTRALE
ARRESTING OFFICER AGENCY: Titusville PD
LAST AGENCY INSPECTION: 06/16/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: HEAD

DATE/TIME OF TEST: 06/26/2017 01:51:56

Diagnostics Check	OK	01:45
Air Blank	0.000	01:46
Control Test	0.079	01:46
Air Blank	0.000	01:46
Subject Sample # 1	0.125	01:47
Breath Volume	2.222	01:47
Air Blank	0.000	01:47
Air Blank	0.000	01:49
Subject Sample # 2	0.126	01:50
Breath Volume	3.460	01:50
Air Blank	0.000	01:51
Control Test	0.080	01:51
Air Blank	0.000	01:51
Diagnostics Check	OK	01:51

Data Download Date/Time: 07/16/2017 12:08