Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,RAY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: ZABALA,JOSE
ARRESTING OFFICER AGENCY: Oviedo PD
LAST AGENCY INSPECTION: 06/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 20:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:25

SUBJECT'S LAST NAME: MAURER

DATE/TIME OF TEST: 07/04/2017 21:52:13

Diagnostics Check		OK	21:45
Air Blank		0.000	21:46
Control Test		0.081	21:46
Air Blank		0.000	21:46
Subject Sample	# 1	0.113	21:47
Breath Volume		2.214	21:47
Air Blank		0.000	21:48
Air Blank		0.000	21:50
Subject Sample	# 2	0.104	21:50
Breath Volume		1.140	21:50
Air Blank		0.000	21:51
Control Test		0.081	21:51
Air Blank		0.000	21:52
Diagnostics Check		OK	21:52

Data Download Date/Time: 07/30/2017 22:10

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,RAY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: TAVANO,MICHAEL
ARRESTING OFFICER AGENCY: FFWCC (Ocala)
LAST AGENCY INSPECTION: 06/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: BUI
ARREST TIME: 23:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: SCHUMACHER

DATE/TIME OF TEST: 07/05/2017 00:57:18

Diagnostics Check		OK	00:50
Air Blank		0.000	00:51
Control Test		0.080	00:51
Air Blank		0.000	00:52
Subject Sample	# 1	0.138	00:52
Breath Volume		3.640	00:52
Air Blank		0.000	00:53
Air Blank		0.000	00:55
Subject Sample	# 2	0.139	00:55
Breath Volume		3.699	00:55
Air Blank		0.000	00:56
Control Test		0.080	00:56
Air Blank		0.000	00:57
Diagnostics Check		OK	00:57

Data Download Date/Time: 07/30/2017 22:10

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,RAY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: NEAL,MARTIN
ARRESTING OFFICER AGENCY: Lake Mary PD
LAST AGENCY INSPECTION: 06/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 23:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: MARTINEZ-HADDEN

DATE/TIME OF TEST: 07/06/2017 02:54:31

Diagnostics Check		OK	02:48
Air Blank		0.000	02:48
Control Test		0.081	02:48
Air Blank		0.000	02:49
Subject Sample	# 1	0.221	02:50
Breath Volume		2.582	02:50
Air Blank		0.000	02:50
Air Blank		0.000	02:52
Subject Sample	# 2	0.220	02:53
Breath Volume		2.273	02:53
Air Blank		0.000	02:53
Control Test		0.080	02:53
Air Blank		0.000	02:54
Diagnostics Check		OK	02:54

Data Download Date/Time: 07/30/2017 22:10

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRANSUE,ZACHARY,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: DRAUGHTON,BOBBY
ARRESTING OFFICER AGENCY: Oviedo PD
LAST AGENCY INSPECTION: 06/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 18:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:48

SUBJECT'S LAST NAME: COOK

DATE/TIME OF TEST: 07/06/2017 20:16:37

Diagnostics Check		OK	20:10
Air Blank		0.000	20:11
Control Test		0.081	20:11
Air Blank		0.000	20:12
Subject Sample	# 1	0.000	20:12
Breath Volume		3.289	20:12
Air Blank		0.000	20:12
Air Blank		0.000	20:14
Subject Sample	# 2	0.000	20:15
Breath Volume		2.312	20:15
Air Blank		0.000	20:15
Control Test		0.080	20:16
Air Blank		0.000	20:16
Diagnostics Check		OK	20:16

Data Download Date/Time: 07/30/2017 22:10

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,RAY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: COLBURN,MEGAN
ARRESTING OFFICER AGENCY: Casselberry PD
LAST AGENCY INSPECTION: 06/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 20:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: MOREFIELD

DATE/TIME OF TEST: 07/09/2017 22:04:27

Diagnostics Check		OK	21:55
Air Blank		0.000	21:56
Control Test		0.081	21:56
Air Blank		0.000	21:56
Subject Sample	# 1	0.068	22:00
Breath Volume		1.554	22:00
Air Blank		0.000	22:00
Air Blank		0.000	22:02
Subject Sample	# 2	0.066	22:02
Breath Volume		1.449	22:02
Air Blank		0.000	22:03
Control Test		0.081	22:03
Air Blank		0.000	22:04
Diagnostics Check		OK	22:04

Data Download Date/Time: 07/30/2017 22:10

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,RAY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: FOWLER,MATTHEW
ARRESTING OFFICER AGENCY: Altamonte Springs PD
LAST AGENCY INSPECTION: 06/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 00:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: MARTINEZ ALVAREZ

DATE/TIME OF TEST: 07/10/2017 01:21:34

Diagnostics Check		OK	01:14
Air Blank		0.000	01:14
Control Test		0.080	01:15
Air Blank		0.000	01:15
Subject Sample	# 1	0.229	01:17
Breath Volume		2.457	01:17
Air Blank		0.000	01:17
Air Blank		0.000	01:19
Subject Sample	# 2	0.226	01:20
Breath Volume		1.500	01:20
Air Blank		0.000	01:20
Control Test		0.080	01:20
Air Blank		0.000	01:21
Diagnostics Check		OK	01:21

Data Download Date/Time: 07/30/2017 22:10

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,RAY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: VAN HEUSEN,ROBERT
ARRESTING OFFICER AGENCY: Oviedo PD
LAST AGENCY INSPECTION: 06/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 00:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: PICKARD

DATE/TIME OF TEST: 07/10/2017 01:49:44

Diagnostics Check		OK	01:43
Air Blank		0.000	01:44
Control Test		0.079	01:44
Air Blank		0.000	01:44
Subject Sample	# 1	0.068	01:45
Breath Volume		2.386	01:45
Air Blank		0.000	01:45
Air Blank		0.000	01:47
Subject Sample	# 2	0.064	01:48
Breath Volume		2.335	01:48
Air Blank		0.000	01:48
Control Test		0.080	01:49
Air Blank		0.000	01:49
Diagnostics Check		OK	01:49

Data Download Date/Time: 07/30/2017 22:10

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMMENS,ADAM,J
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: MEADOWS,SEAN
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 06/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 17:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:08

SUBJECT'S LAST NAME: BORRERO

DATE/TIME OF TEST: 07/13/2017 18:43:28

Diagnostics Check		OK	18:37
Air Blank		0.000	18:37
Control Test		0.080	18:38
Air Blank		0.000	18:38
Subject Sample	# 1	0.100	18:38
Breath Volume		3.574	18:38
Air Blank		0.000	18:39
Air Blank		0.000	18:41
Subject Sample	# 2	0.100	18:41
Breath Volume		3.812	18:41
Air Blank		0.000	18:42
Control Test		0.081	18:42
Air Blank		0.000	18:43
Diagnostics Check		OK	18:43

Data Download Date/Time: 07/30/2017 22:10

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLBURN,MEGAN,E
BREATH TEST OPERATOR AGENCY: Casselberry PD
ARRESTING OFFICER: COLBURN,MEGAN
ARRESTING OFFICER AGENCY: Casselberry PD
LAST AGENCY INSPECTION: 06/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 03:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: NEGRON PEREZ

DATE/TIME OF TEST: 07/14/2017 05:06:20

Diagnostics Check		OK	05:04	
Air Blank		0.000	05:04	
Control Test		0.081	05:04	
Air Blank		0.000	05:05	
Subject Sample	# 1		05:06	RFI Detect
Breath Volume		0.000	05:06	
Air Blank			05:06	RFI Detect
Air Blank			05:06	RFI Detect

Data Download Date/Time: 07/30/2017 22:10

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLBURN,MEGAN,E
BREATH TEST OPERATOR AGENCY: Casselberry PD
ARRESTING OFFICER: COLBURN,MEGAN
ARRESTING OFFICER AGENCY: Casselberry PD
LAST AGENCY INSPECTION: 06/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 03:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: NEGRON PEREZ

DATE/TIME OF TEST: 07/14/2017 05:18:51

Diagnostics Check		OK	05:08	
Air Blank		0.000	05:08	
Control Test		0.081	05:08	
Air Blank		0.000	05:09	
Subject Sample	# 1	0.138	05:12	Volume Not Met
Breath Volume		0.031	05:12	
Air Blank		0.000	05:13	
Air Blank		0.000	05:14	
Subject Sample	# 2	0.129	05:18	Volume Not Met
Breath Volume		0.484	05:18	
Air Blank			05:18	RFI Detect
Air Blank		0.000	05:18	

Data Download Date/Time: 07/30/2017 22:10

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLBURN,MEGAN,E
BREATH TEST OPERATOR AGENCY: Casselberry PD
ARRESTING OFFICER: SEAL,ZACHARY
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 06/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 04:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:15

SUBJECT'S LAST NAME: JEFFERSON

DATE/TIME OF TEST: 07/14/2017 05:50:12

Diagnostics Check		OK	05:43
Air Blank		0.000	05:44
Control Test		0.081	05:44
Air Blank		0.000	05:45
Subject Sample	# 1	0.243	05:45
Breath Volume		1.957	05:45
Air Blank		0.000	05:46
Air Blank		0.000	05:48
Subject Sample	# 2	0.243	05:48
Breath Volume		1.980	05:48
Air Blank		0.000	05:49
Control Test		0.080	05:49
Air Blank		0.000	05:50
Diagnostics Check		OK	05:50

Data Download Date/Time: 07/30/2017 22:10

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,RAY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: TORRES CRUZ,KENNY
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 06/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 20:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: BOVE

DATE/TIME OF TEST: 07/18/2017 21:48:57

Diagnostics Check		OK	21:41
Air Blank		0.000	21:41
Control Test		0.080	21:41
Air Blank		0.000	21:42
Subject Sample	# 1	0.204	21:43
Breath Volume		1.308	21:43
Air Blank		0.000	21:44
Air Blank		0.000	21:46
Subject Sample	# 2	0.203	21:47
Breath Volume		1.492	21:47
Air Blank		0.000	21:48
Control Test		0.081	21:48
Air Blank		0.000	21:48
Diagnostics Check		OK	21:48

Data Download Date/Time: 07/30/2017 22:10

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,RAY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: COREY,TIMOTHY
ARRESTING OFFICER AGENCY: Sanford PD
LAST AGENCY INSPECTION: 06/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 02:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: RAMSDELL

DATE/TIME OF TEST: 07/19/2017 03:05:28

Diagnostics Check		OK	02:59
Air Blank		0.000	02:59
Control Test		0.081	03:00
Air Blank		0.000	03:00
Subject Sample	# 1	0.000	03:00
Breath Volume		1.402	03:00
Air Blank		0.000	03:01
Air Blank		0.000	03:03
Subject Sample	# 2	0.000	03:04
Breath Volume		3.234	03:04
Air Blank		0.000	03:04
Control Test		0.080	03:04
Air Blank		0.000	03:05
Diagnostics Check		OK	03:05

Data Download Date/Time: 07/30/2017 22:10

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,RAY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: COLBURN,MEGAN
ARRESTING OFFICER AGENCY: Casselberry PD
LAST AGENCY INSPECTION: 06/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: BRAVO HERNANDEZ

DATE/TIME OF TEST: 07/24/2017 02:27:24

Diagnostics Check		OK	02:15	
Air Blank		0.000	02:16	
Control Test		0.080	02:16	
Air Blank		0.000	02:17	
Subject Sample	# 1	0.116	02:20	Volume Not Met
Breath Volume		0.324	02:20	
Air Blank		0.000	02:20	
Air Blank		0.000	02:22	
Subject Sample	# 2	0.127	02:25	Volume Not Met
Breath Volume		0.148	02:25	
Air Blank		0.000	02:26	
Control Test		0.081	02:26	
Air Blank		0.000	02:27	
Diagnostics Check		OK	02:27	

Data Download Date/Time: 07/30/2017 22:10

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,RAY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: PRICE,ZACHARY
ARRESTING OFFICER AGENCY: Casselberry PD
LAST AGENCY INSPECTION: 06/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: WALLACE

DATE/TIME OF TEST: 07/24/2017 04:27:44

Diagnostics Check		OK	04:20
Air Blank		0.000	04:21
Control Test		0.080	04:21
Air Blank		0.000	04:22
Subject Sample	# 1	0.000	04:23
Breath Volume		1.812	04:23
Air Blank		0.000	04:24
Air Blank		0.000	04:26
Subject Sample	# 2	0.000	04:26
Breath Volume		1.894	04:26
Air Blank		0.000	04:26
Control Test		0.080	04:27
Air Blank		0.000	04:27
Diagnostics Check		OK	04:27

Data Download Date/Time: 07/30/2017 22:11

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,RAY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: TEDESCO,JOSEPH
ARRESTING OFFICER AGENCY: Oviedo PD
LAST AGENCY INSPECTION: 06/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 17:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: FONT

DATE/TIME OF TEST: 07/24/2017 23:07:41

Diagnostics Check		OK	22:54	
Air Blank		0.000	22:55	
Control Test		0.080	22:55	
Air Blank		0.000	22:55	
Subject Sample	# 1	0.134	22:57	
Breath Volume		1.328	22:57	
Air Blank		0.000	22:57	
Air Blank		0.000	22:59	
Subject Sample	# 2	0.120	23:03	Volume Not Met
Breath Volume		0.726	23:03	
Air Blank		0.000	23:03	
Air Blank		0.000	23:05	
Subject Sample	# 3	0.126	23:06	
Breath Volume		1.437	23:06	
Air Blank		0.000	23:06	
Control Test		0.079	23:07	
Air Blank		0.000	23:07	
Diagnostics Check		OK	23:07	

Data Download Date/Time: 07/30/2017 22:11

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,ANTHONY,M
BREATH TEST OPERATOR AGENCY: Satellite Beach PD
ARRESTING OFFICER: HOLLAND,ANTHONY
ARRESTING OFFICER AGENCY: Satellite Beach PD
LAST AGENCY INSPECTION: 06/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 09:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:25

SUBJECT'S LAST NAME: BIRNBAUM

DATE/TIME OF TEST: 07/11/2017 10:52:18

Diagnostics Check		OK	10:45
Air Blank		0.000	10:45
Control Test		0.079	10:46
Air Blank		0.000	10:46
Subject Sample	# 1	0.000	10:47
Breath Volume		1.496	10:47
Air Blank		0.000	10:47
Air Blank		0.000	10:49
Subject Sample	# 2	0.000	10:50
Breath Volume		1.800	10:50
Air Blank		0.000	10:51
Control Test		0.078	10:51
Air Blank		0.000	10:52
Diagnostics Check		OK	10:52

Data Download Date/Time: 07/27/2017 06:41

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOUSMAN,JOHN,D
BREATH TEST OPERATOR AGENCY: Satellite Beach PD
ARRESTING OFFICER: HOUSMAN,JOHN
ARRESTING OFFICER AGENCY: Satellite Beach PD
LAST AGENCY INSPECTION: 06/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 08:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:35

SUBJECT'S LAST NAME: SOARES

DATE/TIME OF TEST: 07/23/2017 09:53:01

Diagnostics Check		OK	09:46
Air Blank		0.000	09:46
Control Test		0.080	09:47
Air Blank		0.000	09:47
Subject Sample	# 1	0.000	09:48
Breath Volume		1.261	09:48
Air Blank		0.000	09:49
Air Blank		0.000	09:51
Subject Sample	# 2	0.000	09:51
Breath Volume		1.589	09:51
Air Blank		0.000	09:52
Control Test		0.079	09:52
Air Blank		0.000	09:52
Diagnostics Check		OK	09:53

Data Download Date/Time: 07/27/2017 06:41

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001002

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASSICK,PATRICIA,A
BREATH TEST OPERATOR AGENCY: Hardee County SO
ARRESTING OFFICER: SPENCER,ROBERT
ARRESTING OFFICER AGENCY: Wauchula PD
LAST AGENCY INSPECTION: 06/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 00:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:17

SUBJECT'S LAST NAME: LY

DATE/TIME OF TEST: 07/02/2017 01:12:17

Diagnostics Check		OK	01:00	
Air Blank		0.000	01:01	
Control Test		0.078	01:01	
Air Blank		0.000	01:02	
Subject Sample	# 1		01:05	No Sample Provided
Breath Volume		0.000	01:05	
Air Blank		0.000	01:05	
Air Blank		0.000	01:07	
Subject Sample	# 2		01:10	No Sample Provided
Breath Volume		0.000	01:10	
Air Blank		0.000	01:11	
Control Test		0.079	01:11	
Air Blank		0.000	01:12	
Diagnostics Check		OK	01:12	

Data Download Date/Time: 07/26/2017 11:30

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001002

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNEIDER,MARK,A
BREATH TEST OPERATOR AGENCY: Hardee County SO
ARRESTING OFFICER: WYATT,JUSTIN
ARRESTING OFFICER AGENCY: Wauchula PD
LAST AGENCY INSPECTION: 06/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 18:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:29

SUBJECT'S LAST NAME: MOSES

DATE/TIME OF TEST: 07/04/2017 18:59:20

Diagnostics Check		OK	18:56	
Air Blank		0.000	18:56	
Control Test		0.079	18:57	
Air Blank		0.000	18:57	
Subject Sample	# 1		18:57	Subject Test Refused
Breath Volume		0.000	18:57	
Air Blank		0.000	18:58	
Control Test		0.078	18:58	
Air Blank		0.000	18:59	
Diagnostics Check		OK	18:59	

Data Download Date/Time: 07/26/2017 11:30

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001002

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARAUJO,NESTOR,J
BREATH TEST OPERATOR AGENCY: Hardee County SO
ARRESTING OFFICER: WYATT,J
ARRESTING OFFICER AGENCY: Wauchula PD
LAST AGENCY INSPECTION: 06/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 00:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:23

SUBJECT'S LAST NAME: RODRIGUEZ-GARCIA

DATE/TIME OF TEST: 07/10/2017 00:47:29

Diagnostics Check		OK	00:43	
Air Blank		0.000	00:44	
Control Test		0.077	00:44	
Air Blank		0.000	00:45	
Subject Sample	# 1		00:45	Slope Not Met
Breath Volume		1.242	00:45	
Air Blank		0.000	00:46	
Control Test		0.078	00:46	
Air Blank		0.000	00:47	
Diagnostics Check		OK	00:47	

Data Download Date/Time: 07/26/2017 11:30

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001002

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARAUJO,NESTOR,J
BREATH TEST OPERATOR AGENCY: Hardee County SO
ARRESTING OFFICER: WYATT,J
ARRESTING OFFICER AGENCY: Wauchula PD
LAST AGENCY INSPECTION: 06/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 00:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:23

SUBJECT'S LAST NAME: RODRIGUEZ-GARCIA

DATE/TIME OF TEST: 07/10/2017 00:57:54

Diagnostics Check		OK	00:50
Air Blank		0.000	00:51
Control Test		0.079	00:51
Air Blank		0.000	00:51
Subject Sample	# 1	0.286	00:52
Breath Volume		2.054	00:52
Air Blank		0.000	00:53
Air Blank		0.000	00:55
Subject Sample	# 2	0.289	00:56
Breath Volume		1.480	00:56
Air Blank		0.000	00:57
Control Test		0.078	00:57
Air Blank		0.000	00:57
Diagnostics Check		OK	00:57

Data Download Date/Time: 07/26/2017 11:30

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001002

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARAUJO,NESTOR,J
BREATH TEST OPERATOR AGENCY: Hardee County SO
ARRESTING OFFICER: WYATT,J
ARRESTING OFFICER AGENCY: Wauchula PD
LAST AGENCY INSPECTION: 06/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 20:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: ROSALES

DATE/TIME OF TEST: 07/15/2017 21:15:11

Diagnostics Check		OK	21:12	
Air Blank		0.000	21:13	
Control Test		0.080	21:13	
Air Blank		0.000	21:13	
Subject Sample	# 1		21:14	Sequence Aborted
Breath Volume		0.000	21:14	
Air Blank		0.000	21:15	

Data Download Date/Time: 07/26/2017 11:30

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001002

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARAUJO,NESTOR,J
BREATH TEST OPERATOR AGENCY: Hardee County SO
ARRESTING OFFICER: SPENCER,ROBERT
ARRESTING OFFICER AGENCY: Wauchula PD
LAST AGENCY INSPECTION: 06/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 22:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: MANRIQUEZ PAREDES

DATE/TIME OF TEST: 07/20/2017 23:11:30

Diagnostics Check		OK	23:08	
Air Blank		0.000	23:09	
Control Test		0.079	23:09	
Air Blank		0.000	23:09	
Subject Sample	# 1		23:10	Subject Test Refused
Breath Volume		0.000	23:10	
Air Blank		0.000	23:10	
Control Test		0.079	23:10	
Air Blank		0.000	23:11	
Diagnostics Check		OK	23:11	

Data Download Date/Time: 07/26/2017 11:30

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001002

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTWRIGHT JR,GARY,W
BREATH TEST OPERATOR AGENCY: Hardee County SO
ARRESTING OFFICER: LOCKE,BREANNA
ARRESTING OFFICER AGENCY: Bowling Green PD
LAST AGENCY INSPECTION: 06/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 23:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:31

SUBJECT'S LAST NAME: LINKENHOGER

DATE/TIME OF TEST: 07/22/2017 01:03:17

Diagnostics Check		OK	01:00	
Air Blank		0.000	01:00	
Control Test		0.079	01:01	
Air Blank		0.000	01:01	
Subject Sample	# 1		01:01	Subject Test Refused
Breath Volume		0.000	01:01	
Air Blank		0.000	01:02	
Control Test		0.079	01:02	
Air Blank		0.000	01:03	
Diagnostics Check		OK	01:03	

Data Download Date/Time: 07/26/2017 11:30

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KASELAK,STEPHEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 00:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:57

SUBJECT'S LAST NAME: VAUGHN II

DATE/TIME OF TEST: 07/01/2017 01:28:55

Diagnostics Check		OK	01:22
Air Blank		0.000	01:23
Control Test		0.081	01:23
Air Blank		0.000	01:23
Subject Sample	# 1	0.243	01:24
Breath Volume		2.074	01:24
Air Blank		0.000	01:24
Air Blank		0.000	01:26
Subject Sample	# 2	0.240	01:27
Breath Volume		2.042	01:27
Air Blank		0.000	01:28
Control Test		0.080	01:28
Air Blank		0.000	01:28
Diagnostics Check		OK	01:28

Data Download Date/Time: 07/28/2017 21:40

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PERSAUD,COMPTON,D
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: PERSAUD,COMPTON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 19:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:05

SUBJECT'S LAST NAME: SEVIGNY

DATE/TIME OF TEST: 07/01/2017 20:37:38

Diagnostics Check		OK	20:26	
Air Blank		0.000	20:26	
Control Test		0.081	20:26	
Air Blank		0.000	20:27	
Subject Sample	# 1	0.193	20:30	Volume Not Met
Breath Volume		0.695	20:30	
Air Blank		0.000	20:31	
Air Blank		0.000	20:32	
Subject Sample	# 2	0.177	20:36	Volume Not Met
Breath Volume		0.906	20:36	
Air Blank		0.000	20:36	
Control Test		0.080	20:37	
Air Blank		0.000	20:37	
Diagnostics Check		OK	20:37	

Data Download Date/Time: 07/28/2017 21:40

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PERSAUD,COMPTON,D
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: PERSAUD,COMPTON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 19:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:05

SUBJECT'S LAST NAME: SEVIGNY

DATE/TIME OF TEST: 07/01/2017 20:50:10

Diagnostics Check		OK	20:39	
Air Blank		0.000	20:39	
Control Test		0.079	20:39	
Air Blank		0.000	20:40	
Subject Sample	# 1	0.196	20:43	Volume Not Met
Breath Volume		1.042	20:43	
Air Blank		0.000	20:44	
Air Blank		0.000	20:45	
Subject Sample	# 2		20:48	Subject Test Refused
Breath Volume		1.085	20:48	
Air Blank		0.000	20:49	
Control Test		0.079	20:49	
Air Blank		0.000	20:50	
Diagnostics Check		OK	20:50	

Data Download Date/Time: 07/28/2017 21:40

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AMATRUDA,CHRIS,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: AMATRUDA,CHRIS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 22:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:58

SUBJECT'S LAST NAME: EDWIN

DATE/TIME OF TEST: 07/01/2017 23:24:37

Diagnostics Check		OK	23:18
Air Blank		0.000	23:19
Control Test		0.080	23:19
Air Blank		0.000	23:19
Subject Sample	# 1	0.000	23:20
Breath Volume		1.597	23:20
Air Blank		0.000	23:20
Air Blank		0.000	23:22
Subject Sample	# 2	0.000	23:23
Breath Volume		1.257	23:23
Air Blank		0.000	23:23
Control Test		0.081	23:24
Air Blank		0.000	23:24
Diagnostics Check		OK	23:24

Data Download Date/Time: 07/28/2017 21:40

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AMATRUDA,CHRIS,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: AMATRUDA,CHRIS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: WALLS

DATE/TIME OF TEST: 07/02/2017 03:16:27

Diagnostics Check	OK	03:15	
Air Blank		03:15	RFI Detect
Air Blank	0.000	03:16	

Data Download Date/Time: 07/28/2017 21:40

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AMATRUDA,CHRIS,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: AMATRUDA,CHRIS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: WALLS

DATE/TIME OF TEST: 07/02/2017 03:26:12

Diagnostics Check		OK	03:18
Air Blank		0.000	03:19
Control Test		0.080	03:19
Air Blank		0.000	03:20
Subject Sample	# 1	0.176	03:21
Breath Volume		2.148	03:21
Air Blank		0.000	03:22
Air Blank		0.000	03:23
Subject Sample	# 2	0.180	03:24
Breath Volume		1.332	03:24
Air Blank		0.000	03:25
Control Test		0.081	03:25
Air Blank		0.000	03:26
Diagnostics Check		OK	03:26

Data Download Date/Time: 07/28/2017 21:40

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORRIS,BRIDGETT,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: FORD,ERIC
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 22:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: NELSON

DATE/TIME OF TEST: 07/02/2017 23:49:17

Diagnostics Check		OK	23:40	
Air Blank		0.000	23:41	
Control Test		0.080	23:41	
Air Blank		0.000	23:42	
Subject Sample	# 1		23:45	No Sample Provided
Breath Volume		0.000	23:45	
Air Blank		0.000	23:45	
Air Blank		0.000	23:47	
Subject Sample	# 2		23:47	Subject Test Refused
Breath Volume		0.000	23:47	
Air Blank		0.000	23:48	
Control Test		0.081	23:48	
Air Blank		0.000	23:49	
Diagnostics Check		OK	23:49	

Data Download Date/Time: 07/28/2017 21:41

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HORNING,PETER,T
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: FOSTER,ALEXANDER
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 22:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:59

SUBJECT'S LAST NAME: SCOTT

DATE/TIME OF TEST: 07/03/2017 23:22:15

Diagnostics Check		OK	23:19	
Air Blank		0.000	23:19	
Control Test		0.081	23:20	
Air Blank		0.000	23:20	
Subject Sample	# 1		23:20	Subject Test Refused
Breath Volume		0.000	23:20	
Air Blank		0.000	23:21	
Control Test		0.080	23:21	
Air Blank		0.000	23:22	
Diagnostics Check		OK	23:22	

Data Download Date/Time: 07/28/2017 21:41

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HORNING,PETER,T
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: CAREY,KEVIN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 23:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:39

SUBJECT'S LAST NAME: HEHENBERGER

DATE/TIME OF TEST: 07/04/2017 01:02:22

Diagnostics Check		OK	00:59	
Air Blank		0.000	01:00	
Control Test		0.081	01:00	
Air Blank		0.000	01:00	
Subject Sample	# 1		01:01	Subject Test Refused
Breath Volume		0.000	01:01	
Air Blank		0.000	01:01	
Control Test		0.080	01:01	
Air Blank		0.000	01:02	
Diagnostics Check		OK	01:02	

Data Download Date/Time: 07/28/2017 21:41

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HORNING,PETER,T
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HANNA,JANE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 00:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:02

SUBJECT'S LAST NAME: BERMUDEZ

DATE/TIME OF TEST: 07/04/2017 01:27:37

Diagnostics Check		OK	01:22	
Air Blank		0.000	01:23	
Control Test		0.080	01:23	
Air Blank		0.000	01:24	
Subject Sample	# 1		01:26	Subject Test Refused
Breath Volume		0.000	01:26	
Air Blank		0.000	01:26	
Control Test		0.081	01:27	
Air Blank		0.000	01:27	
Diagnostics Check		OK	01:27	

Data Download Date/Time: 07/28/2017 21:41

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HORNING,PETER,T
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MULLINS,JOHN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:28

SUBJECT'S LAST NAME: PORCELLA

DATE/TIME OF TEST: 07/04/2017 02:04:03

Diagnostics Check		OK	01:52	
Air Blank		0.000	01:52	
Control Test		0.081	01:53	
Air Blank		0.000	01:53	
Subject Sample	# 1	0.097	01:56	Volume Not Met
Breath Volume		0.117	01:56	
Air Blank		0.000	01:57	
Air Blank		0.000	01:59	
Subject Sample	# 2	0.126	02:02	Volume Not Met
Breath Volume		0.421	02:02	
Air Blank		0.000	02:03	
Control Test		0.081	02:03	
Air Blank		0.000	02:03	
Diagnostics Check		OK	02:04	

Data Download Date/Time: 07/28/2017 21:41

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HORNING,PETER,T
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MULLINS,JOHN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:28

SUBJECT'S LAST NAME: PORCELLA

DATE/TIME OF TEST: 07/04/2017 02:16:48

Diagnostics Check		OK	02:08	
Air Blank		0.000	02:08	
Control Test		0.080	02:09	
Air Blank		0.000	02:09	
Subject Sample	# 1	0.114	02:12	Volume Not Met
Breath Volume		0.308	02:12	
Air Blank		0.000	02:13	
Air Blank		0.000	02:15	
Subject Sample	# 2		02:15	Subject Test Refused
Breath Volume		0.000	02:15	
Air Blank		0.000	02:15	
Control Test		0.079	02:16	
Air Blank		0.000	02:16	
Diagnostics Check		OK	02:16	

Data Download Date/Time: 07/28/2017 21:41

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERGER,CHARLES,K
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BERGER,CHARLES
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 04:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: SCHELL

DATE/TIME OF TEST: 07/04/2017 04:56:55

Diagnostics Check		OK	04:51	
Air Blank		0.000	04:51	
Control Test		0.080	04:51	
Air Blank		0.000	04:52	
Subject Sample	# 1		04:55	Subject Test Refused
Breath Volume		0.000	04:55	
Air Blank		0.000	04:55	
Control Test		0.080	04:55	
Air Blank			04:56	RFI Detect
Air Blank		0.000	04:56	

Data Download Date/Time: 07/28/2017 21:41

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PERSAUD,COMPTON,D
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: PERSAUD,COMPTON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 19:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:21

SUBJECT'S LAST NAME: GRIMM

DATE/TIME OF TEST: 07/04/2017 19:52:52

Diagnostics Check		OK	19:41	
Air Blank		0.000	19:41	
Control Test		0.081	19:42	
Air Blank		0.000	19:42	
Subject Sample	# 1		19:45	No Sample Provided
Breath Volume		0.000	19:45	
Air Blank		0.000	19:46	
Air Blank		0.000	19:48	
Subject Sample	# 2		19:51	No Sample Provided
Breath Volume		0.000	19:51	
Air Blank		0.000	19:51	
Control Test		0.080	19:52	
Air Blank		0.000	19:52	
Diagnostics Check		OK	19:52	

Data Download Date/Time: 07/28/2017 21:41

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PERSAUD,COMPTON,D
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: DONOHUE,S
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 19:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:59

SUBJECT'S LAST NAME: MILANA

DATE/TIME OF TEST: 07/04/2017 20:30:55

Diagnostics Check		OK	20:19	
Air Blank		0.000	20:20	
Control Test		0.081	20:20	
Air Blank		0.000	20:20	
Subject Sample	# 1		20:23	No Sample Provided
Breath Volume		0.000	20:23	
Air Blank		0.000	20:24	
Air Blank		0.000	20:26	
Subject Sample	# 2		20:29	No Sample Provided
Breath Volume		0.000	20:29	
Air Blank		0.000	20:30	
Control Test		0.080	20:30	
Air Blank		0.000	20:30	
Diagnostics Check		OK	20:30	

Data Download Date/Time: 07/28/2017 21:41

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KASELAK,STEPHEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 23:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: JEWELL

DATE/TIME OF TEST: 07/05/2017 00:08:13

Diagnostics Check		OK	00:05	
Air Blank		0.000	00:05	
Control Test		0.081	00:06	
Air Blank		0.000	00:06	
Subject Sample	# 1		00:06	Subject Test Refused
Breath Volume		0.000	00:06	
Air Blank		0.000	00:07	
Control Test		0.081	00:07	
Air Blank		0.000	00:08	
Diagnostics Check		OK	00:08	

Data Download Date/Time: 07/28/2017 21:41

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE,JONATHAN,S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WHITEHURST,WYETH
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 02:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:12

SUBJECT'S LAST NAME: MADERA

DATE/TIME OF TEST: 07/05/2017 03:39:14

Diagnostics Check		OK	03:32
Air Blank		0.000	03:32
Control Test		0.080	03:33
Air Blank		0.000	03:33
Subject Sample	# 1	0.052	03:34
Breath Volume		1.398	03:34
Air Blank		0.000	03:35
Air Blank		0.000	03:37
Subject Sample	# 2	0.054	03:37
Breath Volume		1.843	03:37
Air Blank		0.000	03:38
Control Test		0.080	03:38
Air Blank		0.000	03:39
Diagnostics Check		OK	03:39

Data Download Date/Time: 07/28/2017 21:41

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PERSAUD,COMPTON,D
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: PERSAUD,COMPTON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 01:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:53

SUBJECT'S LAST NAME: WHITTENBERG

DATE/TIME OF TEST: 07/06/2017 02:20:26

Diagnostics Check		OK	02:13
Air Blank		0.000	02:13
Control Test		0.080	02:14
Air Blank		0.000	02:14
Subject Sample	# 1	0.185	02:16
Breath Volume		1.605	02:16
Air Blank		0.000	02:16
Air Blank		0.000	02:18
Subject Sample	# 2	0.182	02:18
Breath Volume		1.527	02:18
Air Blank		0.000	02:19
Control Test		0.079	02:19
Air Blank		0.000	02:20
Diagnostics Check		OK	02:20

Data Download Date/Time: 07/28/2017 21:41

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PERSAUD,COMPTON,D
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: PERSAUD,COMPTON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 23:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:24

SUBJECT'S LAST NAME: HERRERAVELAZQUEZ

DATE/TIME OF TEST: 07/07/2017 00:50:52

Diagnostics Check		OK	00:44
Air Blank		0.000	00:45
Control Test		0.080	00:45
Air Blank		0.000	00:45
Subject Sample	# 1	0.230	00:46
Breath Volume		1.703	00:46
Air Blank		0.000	00:46
Air Blank		0.000	00:48
Subject Sample	# 2	0.231	00:49
Breath Volume		2.332	00:49
Air Blank		0.000	00:49
Control Test		0.080	00:50
Air Blank		0.000	00:50
Diagnostics Check		OK	00:50

Data Download Date/Time: 07/28/2017 21:41

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KASELAK,STEPHEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 01:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:59

SUBJECT'S LAST NAME: YOUNCE

DATE/TIME OF TEST: 07/07/2017 02:27:01

Diagnostics Check		OK	02:20
Air Blank		0.000	02:20
Control Test		0.080	02:20
Air Blank		0.000	02:21
Subject Sample	# 1	0.082	02:21
Breath Volume		1.683	02:21
Air Blank		0.000	02:22
Air Blank		0.000	02:24
Subject Sample	# 2	0.082	02:25
Breath Volume		1.789	02:25
Air Blank		0.000	02:26
Control Test		0.081	02:26
Air Blank		0.000	02:26
Diagnostics Check		OK	02:27

Data Download Date/Time: 07/28/2017 21:41

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PERSAUD,COMPTON,D
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: PERSAUD,COMPTON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 01:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: MONROE

DATE/TIME OF TEST: 07/08/2017 02:09:05

Diagnostics Check		OK	02:00
Air Blank		0.000	02:00
Control Test		0.081	02:01
Air Blank		0.000	02:01
Subject Sample	# 1	0.192	02:04
Breath Volume		1.562	02:04
Air Blank		0.000	02:05
Air Blank		0.000	02:07
Subject Sample	# 2	0.175	02:07
Breath Volume		1.488	02:07
Air Blank		0.000	02:08
Control Test		0.079	02:08
Air Blank		0.000	02:09
Diagnostics Check		OK	02:09

Data Download Date/Time: 07/28/2017 21:41

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KASELAK,STEPHEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 01:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:11

SUBJECT'S LAST NAME: HERNANDEZ PADRON

DATE/TIME OF TEST: 07/08/2017 02:42:58

Diagnostics Check		OK	02:36
Air Blank		0.000	02:37
Control Test		0.080	02:37
Air Blank		0.000	02:38
Subject Sample	# 1	0.030	02:38
Breath Volume		1.871	02:38
Air Blank		0.000	02:39
Air Blank		0.000	02:41
Subject Sample	# 2	0.029	02:41
Breath Volume		1.527	02:41
Air Blank		0.000	02:42
Control Test		0.081	02:42
Air Blank		0.000	02:42
Diagnostics Check		OK	02:42

Data Download Date/Time: 07/28/2017 21:41

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PERSAUD,COMPTON,D
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: PERSAUD,COMPTON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 03:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:51

SUBJECT'S LAST NAME: TURNER

DATE/TIME OF TEST: 07/08/2017 04:18:05

Diagnostics Check		OK	04:11
Air Blank		0.000	04:12
Control Test		0.080	04:12
Air Blank		0.000	04:13
Subject Sample	# 1	0.157	04:13
Breath Volume		1.933	04:13
Air Blank		0.000	04:14
Air Blank		0.000	04:16
Subject Sample	# 2	0.158	04:16
Breath Volume		1.820	04:16
Air Blank		0.000	04:17
Control Test		0.080	04:17
Air Blank		0.000	04:17
Diagnostics Check		OK	04:18

Data Download Date/Time: 07/28/2017 21:41

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERGER,CHARLES,K
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BERGER,CHARLES
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 19:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:40

SUBJECT'S LAST NAME: THORNTON

DATE/TIME OF TEST: 07/08/2017 20:08:51

Diagnostics Check		OK	20:00
Air Blank		0.000	20:01
Control Test		0.080	20:01
Air Blank		0.000	20:01
Subject Sample	# 1	0.171	20:02
Breath Volume		1.425	20:02
Air Blank		0.000	20:03
Air Blank		0.000	20:05
Subject Sample	# 2	0.172	20:07
Breath Volume		1.742	20:07
Air Blank		0.000	20:07
Control Test		0.080	20:08
Air Blank		0.000	20:08
Diagnostics Check		OK	20:08

Data Download Date/Time: 07/28/2017 21:41

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WELSHANS,DEAN,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TATE,GREGORY
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 22:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:22

SUBJECT'S LAST NAME: CASTLEBERRY

DATE/TIME OF TEST: 07/12/2017 00:00:04

Diagnostics Check		OK	23:53
Air Blank		0.000	23:53
Control Test		0.081	23:53
Air Blank		0.000	23:54
Subject Sample	# 1	0.078	23:55
Breath Volume		2.156	23:55
Air Blank		0.000	23:56
Air Blank		0.000	23:58
Subject Sample	# 2	0.075	23:58
Breath Volume		1.605	23:58
Air Blank		0.000	23:59
Control Test		0.081	23:59
Air Blank		0.000	23:59
Diagnostics Check		OK	00:00

Data Download Date/Time: 07/28/2017 21:41

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PERSAUD,COMPTON,D
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: PERSAUD,COMPTON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 22:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:57

SUBJECT'S LAST NAME: RAPISALDO

DATE/TIME OF TEST: 07/12/2017 23:32:37

Diagnostics Check		OK	23:17	
Air Blank		0.000	23:18	
Control Test		0.080	23:18	
Air Blank		0.000	23:19	
Subject Sample	# 1	0.189	23:22	Volume Not Met
Breath Volume		0.496	23:22	
Air Blank		0.000	23:23	
Air Blank		0.000	23:24	
Subject Sample	# 2	0.210	23:25	
Breath Volume		1.210	23:25	
Air Blank		0.000	23:26	
Air Blank		0.000	23:27	
Subject Sample	# 3	0.196	23:31	Volume Not Met
Breath Volume		0.855	23:31	
Air Blank		0.000	23:31	
Control Test		0.080	23:32	
Air Blank		0.000	23:32	
Diagnostics Check		OK	23:32	

Data Download Date/Time: 07/28/2017 21:41

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KASELAK,STEPHEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 02:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:48

SUBJECT'S LAST NAME: SCUDDY

DATE/TIME OF TEST: 07/13/2017 03:13:43

Diagnostics Check		OK	03:10	
Air Blank		0.000	03:11	
Control Test		0.080	03:11	
Air Blank		0.000	03:12	
Subject Sample	# 1		03:12	Subject Test Refused
Breath Volume		0.000	03:12	
Air Blank		0.000	03:12	
Control Test		0.081	03:13	
Air Blank		0.000	03:13	
Diagnostics Check		OK	03:13	

Data Download Date/Time: 07/28/2017 21:41

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KASELAK,STEPHEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 21:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:22

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 07/13/2017 22:54:00

Diagnostics Check		OK	22:47
Air Blank		0.000	22:48
Control Test		0.080	22:48
Air Blank		0.000	22:49
Subject Sample	# 1	0.180	22:49
Breath Volume		1.777	22:49
Air Blank		0.000	22:50
Air Blank		0.000	22:52
Subject Sample	# 2	0.172	22:52
Breath Volume		1.925	22:52
Air Blank		0.000	22:53
Control Test		0.080	22:53
Air Blank		0.000	22:53
Diagnostics Check		OK	22:54

Data Download Date/Time: 07/28/2017 21:41

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAIR,RONALD,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAIR,RONALD
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 02:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:28

SUBJECT'S LAST NAME: TRUMAN

DATE/TIME OF TEST: 07/14/2017 03:06:28

Diagnostics Check		OK	03:00
Air Blank		0.000	03:00
Control Test		0.081	03:01
Air Blank		0.000	03:01
Subject Sample	# 1	0.139	03:02
Breath Volume		1.796	03:02
Air Blank		0.000	03:02
Air Blank		0.000	03:04
Subject Sample	# 2	0.134	03:04
Breath Volume		2.039	03:04
Air Blank		0.000	03:05
Control Test		0.080	03:05
Air Blank		0.000	03:06
Diagnostics Check		OK	03:06

Data Download Date/Time: 07/28/2017 21:41

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LANEY,DAMON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 03:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:58

SUBJECT'S LAST NAME: KALKMAN

DATE/TIME OF TEST: 07/14/2017 04:33:00

Diagnostics Check		OK	04:23
Air Blank		0.000	04:23
Control Test		0.081	04:24
Air Blank		0.000	04:24
Subject Sample	# 1	0.122	04:27
Breath Volume		1.457	04:27
Air Blank		0.000	04:28
Air Blank		0.000	04:30
Subject Sample	# 2	0.111	04:31
Breath Volume		1.386	04:31
Air Blank		0.000	04:32
Control Test		0.080	04:32
Air Blank		0.000	04:32
Diagnostics Check		OK	04:33

Data Download Date/Time: 07/28/2017 21:41

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LANEY,DAMON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 00:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:04

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 07/15/2017 01:35:53

Diagnostics Check		OK	01:29
Air Blank		0.000	01:30
Control Test		0.080	01:30
Air Blank		0.000	01:31
Subject Sample	# 1	0.029	01:31
Breath Volume		2.835	01:31
Air Blank		0.000	01:32
Air Blank		0.000	01:34
Subject Sample	# 2	0.029	01:34
Breath Volume		2.253	01:34
Air Blank		0.000	01:34
Control Test		0.080	01:35
Air Blank		0.000	01:35
Diagnostics Check		OK	01:35

Data Download Date/Time: 07/28/2017 21:41

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KASELAK,STEPHEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 03:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:38

SUBJECT'S LAST NAME: MASTIN

DATE/TIME OF TEST: 07/15/2017 04:09:20

Diagnostics Check		OK	03:59
Air Blank		0.000	03:59
Control Test		0.080	04:00
Air Blank		0.000	04:00
Subject Sample	# 1	0.206	04:03
Breath Volume		1.394	04:03
Air Blank		0.000	04:03
Air Blank		0.000	04:05
Subject Sample	# 2	0.194	04:07
Breath Volume		1.722	04:07
Air Blank		0.000	04:08
Control Test		0.080	04:08
Air Blank		0.000	04:09
Diagnostics Check		OK	04:09

Data Download Date/Time: 07/28/2017 21:41

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAIR,RONALD,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAIR,RONALD
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 00:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:02

SUBJECT'S LAST NAME: BENJAMIN

DATE/TIME OF TEST: 07/16/2017 01:32:17

Diagnostics Check		OK	01:24
Air Blank		0.000	01:25
Control Test		0.080	01:25
Air Blank		0.000	01:26
Subject Sample	# 1	0.034	01:27
Breath Volume		1.496	01:27
Air Blank		0.000	01:28
Air Blank		0.000	01:30
Subject Sample	# 2	0.031	01:30
Breath Volume		1.195	01:30
Air Blank		0.000	01:31
Control Test		0.081	01:31
Air Blank		0.000	01:32
Diagnostics Check		OK	01:32

Data Download Date/Time: 07/28/2017 21:41

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PERSAUD,COMPTON,D
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: PERSAUD,COMPTON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 03:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:13

SUBJECT'S LAST NAME: NEZIC

DATE/TIME OF TEST: 07/16/2017 03:44:54

Diagnostics Check		OK	03:33	
Air Blank		0.000	03:33	
Control Test		0.081	03:34	
Air Blank		0.000	03:34	
Subject Sample	# 1	0.077	03:37	Volume Not Met
Breath Volume		0.128	03:37	
Air Blank		0.000	03:38	
Air Blank		0.000	03:40	
Subject Sample	# 2	0.062	03:43	Volume Not Met
Breath Volume		0.558	03:43	
Air Blank		0.000	03:44	
Control Test		0.081	03:44	
Air Blank		0.000	03:44	
Diagnostics Check		OK	03:44	

Data Download Date/Time: 07/28/2017 21:41

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KASELAK,STEPHEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 03:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:47

SUBJECT'S LAST NAME: RINGGOLD

DATE/TIME OF TEST: 07/16/2017 04:12:22

Diagnostics Check		OK	04:09	
Air Blank		0.000	04:09	
Control Test		0.080	04:09	
Air Blank		0.000	04:10	
Subject Sample	# 1		04:11	Subject Test Refused
Breath Volume		0.000	04:11	
Air Blank		0.000	04:11	
Control Test		0.081	04:11	
Air Blank		0.000	04:12	
Diagnostics Check		OK	04:12	

Data Download Date/Time: 07/28/2017 21:41

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE,JONATHAN,S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WHITEHURST,WYETH
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 22:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: MCKAY

DATE/TIME OF TEST: 07/18/2017 23:40:06

Diagnostics Check		OK	23:31	
Air Blank		0.000	23:31	
Control Test		0.080	23:32	
Air Blank		0.000	23:32	
Subject Sample	# 1		23:35	No Sample Provided
Breath Volume		0.000	23:35	
Air Blank		0.000	23:36	
Air Blank		0.000	23:38	
Subject Sample	# 2		23:38	Subject Test Refused
Breath Volume		0.000	23:38	
Air Blank		0.000	23:39	
Control Test		0.081	23:39	
Air Blank		0.000	23:40	
Diagnostics Check		OK	23:40	

Data Download Date/Time: 07/28/2017 21:41

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE,JONATHAN,S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: CAREY,KEVIN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 02:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:18

SUBJECT'S LAST NAME: NGUYEN

DATE/TIME OF TEST: 07/19/2017 03:49:38

Diagnostics Check		OK	03:38	
Air Blank		0.000	03:38	
Control Test		0.081	03:39	
Air Blank		0.000	03:39	
Subject Sample	# 1		03:42	No Sample Provided
Breath Volume		0.000	03:42	
Air Blank		0.000	03:43	
Air Blank		0.000	03:45	
Subject Sample	# 2		03:48	No Sample Provided
Breath Volume		0.000	03:48	
Air Blank		0.000	03:48	
Control Test		0.081	03:49	
Air Blank		0.000	03:49	
Diagnostics Check		OK	03:49	

Data Download Date/Time: 07/28/2017 21:41

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNAVELY,KENT,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LOSTRAGLIO,COREY
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 09:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:19

SUBJECT'S LAST NAME: COPELAND

DATE/TIME OF TEST: 07/21/2017 10:46:27

Diagnostics Check		OK	10:40
Air Blank		0.000	10:40
Control Test		0.081	10:40
Air Blank		0.000	10:41
Subject Sample	# 1	0.000	10:41
Breath Volume		2.144	10:41
Air Blank		0.000	10:42
Air Blank		0.000	10:44
Subject Sample	# 2	0.000	10:45
Breath Volume		1.335	10:45
Air Blank		0.000	10:45
Control Test		0.081	10:45
Air Blank		0.000	10:46
Diagnostics Check		OK	10:46

Data Download Date/Time: 07/28/2017 21:41

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AMATRUDA,CHRIS,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: AMATRUDA,CHRIS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 00:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:07

SUBJECT'S LAST NAME: HART

DATE/TIME OF TEST: 07/22/2017 01:38:42

Diagnostics Check		OK	01:35	
Air Blank		0.000	01:36	
Control Test		0.081	01:36	
Air Blank		0.000	01:37	
Subject Sample	# 1		01:37	Subject Test Refused
Breath Volume		0.000	01:37	
Air Blank		0.000	01:37	
Control Test		0.081	01:38	
Air Blank		0.000	01:38	
Diagnostics Check		OK	01:38	

Data Download Date/Time: 07/28/2017 21:41

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUDBRINK,BRIAN,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: SUDBRINK,BRIAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:53

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 07/22/2017 03:20:40

Diagnostics Check		OK	03:14	
Air Blank		0.000	03:14	
Control Test		0.081	03:14	
Air Blank		0.000	03:15	
Subject Sample	# 1	0.243	03:16	
Breath Volume		1.191	03:16	
Air Blank		0.000	03:17	
Air Blank		0.000	03:18	
Subject Sample	# 2		03:19	Slope Not Met
Breath Volume		0.671	03:19	
Air Blank		0.000	03:19	
Control Test		0.080	03:20	
Air Blank		0.000	03:20	
Diagnostics Check		OK	03:20	

Data Download Date/Time: 07/28/2017 21:41

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUDBRINK,BRIAN,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: SUDBRINK,BRIAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:53

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 07/22/2017 03:28:56

Diagnostics Check		OK	03:22
Air Blank		0.000	03:22
Control Test		0.078	03:23
Air Blank		0.000	03:23
Subject Sample	# 1	0.230	03:24
Breath Volume		1.167	03:24
Air Blank		0.000	03:24
Air Blank		0.000	03:26
Subject Sample	# 2	0.242	03:27
Breath Volume		1.199	03:27
Air Blank		0.000	03:28
Control Test		0.080	03:28
Air Blank		0.000	03:28
Diagnostics Check		OK	03:28

Data Download Date/Time: 07/28/2017 21:41

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BEELER,BRYAN,C
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: BEELER,BRYAN
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 06/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 15:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:15

SUBJECT'S LAST NAME: BAILEY

DATE/TIME OF TEST: 07/22/2017 16:38:00

Diagnostics Check		OK	16:35	
Air Blank		0.000	16:35	
Control Test		0.080	16:35	
Air Blank		0.000	16:36	
Subject Sample	# 1		16:36	Subject Test Refused
Breath Volume		0.000	16:36	
Air Blank		0.000	16:37	
Control Test		0.081	16:37	
Air Blank		0.000	16:37	
Diagnostics Check		OK	16:38	

Data Download Date/Time: 07/28/2017 21:41

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AMATRUDA,CHRIS,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: AMATRUDA,CHRIS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 23:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:22

SUBJECT'S LAST NAME: DEVANEY

DATE/TIME OF TEST: 07/23/2017 00:57:34

Diagnostics Check		OK	00:42	
Air Blank		0.000	00:42	
Control Test		0.080	00:43	
Air Blank		0.000	00:43	
Subject Sample	# 1	0.176	00:44	
Breath Volume		1.367	00:44	
Air Blank		0.000	00:45	
Air Blank		0.000	00:47	
Subject Sample	# 2	0.165	00:50	Volume Not Met
Breath Volume		0.726	00:50	
Air Blank		0.000	00:51	
Air Blank		0.000	00:52	
Subject Sample	# 3	0.165	00:56	Volume Not Met
Breath Volume		0.117	00:56	
Air Blank		0.000	00:56	
Control Test		0.079	00:56	
Air Blank		0.000	00:57	
Diagnostics Check		OK	00:57	

Data Download Date/Time: 07/28/2017 21:41

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AMATRUDA,CHRIS,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: AMATRUDA,CHRIS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 23:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:22

SUBJECT'S LAST NAME: DEVANEY

DATE/TIME OF TEST: 07/23/2017 01:15:53

Diagnostics Check		OK	00:58	
Air Blank		0.000	00:59	
Control Test		0.080	00:59	
Air Blank		0.000	01:00	
Subject Sample	# 1	0.169	01:03	
Breath Volume		1.437	01:03	
Air Blank		0.000	01:03	
Air Blank		0.000	01:05	
Subject Sample	# 2	0.158	01:08	Volume Not Met
Breath Volume		0.304	01:08	
Air Blank		0.000	01:09	
Air Blank		0.000	01:11	
Subject Sample	# 3	0.131	01:14	Volume Not Met
Breath Volume		0.835	01:14	
Air Blank		0.000	01:14	
Control Test		0.079	01:15	
Air Blank		0.000	01:15	
Diagnostics Check		OK	01:15	

Data Download Date/Time: 07/28/2017 21:41

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AMATRUDA,CHRIS,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: AMATRUDA,CHRIS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 23:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:22

SUBJECT'S LAST NAME: DEVANEY

DATE/TIME OF TEST: 07/23/2017 01:21:18

Diagnostics Check		OK	01:17	
Air Blank		0.000	01:17	
Control Test		0.080	01:17	
Air Blank		0.000	01:18	
Subject Sample	# 1		01:19	Subject Test Refused
Breath Volume		0.484	01:19	
Air Blank		0.000	01:20	
Control Test		0.080	01:20	
Air Blank		0.000	01:21	
Diagnostics Check		OK	01:21	

Data Download Date/Time: 07/28/2017 21:41

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRIFFIN,KYLE,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GRIFFIN,KYLE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 05:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:50

SUBJECT'S LAST NAME: SHOBE

DATE/TIME OF TEST: 07/23/2017 06:18:32

Diagnostics Check		OK	06:12
Air Blank		0.000	06:12
Control Test		0.080	06:13
Air Blank		0.000	06:13
Subject Sample	# 1	0.102	06:14
Breath Volume		1.859	06:14
Air Blank		0.000	06:14
Air Blank		0.000	06:16
Subject Sample	# 2	0.096	06:17
Breath Volume		1.371	06:17
Air Blank		0.000	06:17
Control Test		0.081	06:17
Air Blank		0.000	06:18
Diagnostics Check		OK	06:18

Data Download Date/Time: 07/28/2017 21:41

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SOMADO,VIEGNAME,M
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: REFUSE,RONY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:05

SUBJECT'S LAST NAME: SANCHEZ

DATE/TIME OF TEST: 07/23/2017 23:37:41

Diagnostics Check		OK	23:30
Air Blank		0.000	23:31
Control Test		0.080	23:31
Air Blank		0.000	23:31
Subject Sample	# 1	0.163	23:32
Breath Volume		1.406	23:32
Air Blank		0.000	23:33
Air Blank		0.000	23:35
Subject Sample	# 2	0.166	23:36
Breath Volume		2.410	23:36
Air Blank		0.000	23:36
Control Test		0.080	23:37
Air Blank		0.000	23:37
Diagnostics Check		OK	23:37

Data Download Date/Time: 07/28/2017 21:41

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SOMADO,VIEGNAME,M
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: REFUSE,RONY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 02:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: RAMOS III

DATE/TIME OF TEST: 07/25/2017 03:51:58

Diagnostics Check		OK	03:45
Air Blank		0.000	03:45
Control Test		0.081	03:46
Air Blank		0.000	03:46
Subject Sample	# 1	0.269	03:47
Breath Volume		1.835	03:47
Air Blank		0.000	03:48
Air Blank		0.000	03:49
Subject Sample	# 2	0.258	03:50
Breath Volume		1.550	03:50
Air Blank		0.000	03:51
Control Test		0.079	03:51
Air Blank		0.000	03:51
Diagnostics Check		OK	03:51

Data Download Date/Time: 07/28/2017 21:41

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUCKERS,ROBERT,S
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: DUCKERS,ROBERT
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 19:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: HORAN

DATE/TIME OF TEST: 07/25/2017 21:29:20

Diagnostics Check		OK	21:26	
Air Blank		0.000	21:27	
Control Test		0.081	21:27	
Air Blank		0.000	21:27	
Subject Sample	# 1		21:27	Subject Test Refused
Breath Volume		0.000	21:27	
Air Blank		0.000	21:28	
Control Test		0.082	21:28	
Air Blank		0.000	21:29	
Diagnostics Check		OK	21:29	

Data Download Date/Time: 07/28/2017 21:41

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACOBY,LARRY,K
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: JACOBY,LARRY
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 22:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:37

SUBJECT'S LAST NAME: FERREIRA

DATE/TIME OF TEST: 07/25/2017 23:03:23

Diagnostics Check		OK	22:58	
Air Blank		0.000	22:58	
Control Test		0.082	22:59	
Air Blank		0.000	22:59	
Subject Sample	# 1		23:02	Subject Test Refused
Breath Volume		0.000	23:02	
Air Blank		0.000	23:02	
Control Test		0.080	23:02	
Air Blank		0.000	23:03	
Diagnostics Check		OK	23:03	

Data Download Date/Time: 07/28/2017 21:41

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERGER,CHARLES,K
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BERGER,CHARLES
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 22:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: STARK

DATE/TIME OF TEST: 07/27/2017 23:12:06

Diagnostics Check		OK	23:01	
Air Blank		0.000	23:01	
Control Test		0.080	23:01	
Air Blank		0.000	23:02	
Subject Sample	# 1		23:05	No Sample Provided
Breath Volume		0.000	23:05	
Air Blank		0.000	23:06	
Air Blank		0.000	23:07	
Subject Sample	# 2		23:10	Subject Test Refused
Breath Volume		0.640	23:10	
Air Blank		0.000	23:11	
Control Test		0.079	23:11	
Air Blank		0.000	23:12	
Diagnostics Check		OK	23:12	

Data Download Date/Time: 07/28/2017 21:41

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERGER,CHARLES,K
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BERGER,CHARLES
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 22:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: STARK

DATE/TIME OF TEST: 07/27/2017 23:23:13

Diagnostics Check		OK	23:14
Air Blank		0.000	23:15
Control Test		0.079	23:15
Air Blank		0.000	23:16
Subject Sample	# 1	0.224	23:18
Breath Volume		1.246	23:18
Air Blank		0.000	23:18
Air Blank		0.000	23:20
Subject Sample	# 2	0.219	23:21
Breath Volume		1.230	23:21
Air Blank		0.000	23:22
Control Test		0.080	23:22
Air Blank		0.000	23:23
Diagnostics Check		OK	23:23

Data Download Date/Time: 07/28/2017 21:41

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORRIS,BRIDGETT,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MORRIS,BRIDGETT
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 03:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:14

SUBJECT'S LAST NAME: VIARS

DATE/TIME OF TEST: 07/30/2017 04:37:12

Diagnostics Check		OK	04:34	
Air Blank		0.000	04:34	
Control Test		0.081	04:35	
Air Blank		0.000	04:35	
Subject Sample	# 1		04:35	Subject Test Refused
Breath Volume		0.000	04:35	
Air Blank		0.000	04:36	
Control Test		0.081	04:36	
Air Blank		0.000	04:37	
Diagnostics Check		OK	04:37	

Data Download Date/Time: 08/10/2017 21:53

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORRIS,BRIDGETT,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: SKALKO,RYAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:24

SUBJECT'S LAST NAME: MCCLAIN

DATE/TIME OF TEST: 07/31/2017 01:50:54

Diagnostics Check		OK	01:44
Air Blank		0.000	01:45
Control Test		0.080	01:45
Air Blank		0.000	01:45
Subject Sample	# 1	0.174	01:46
Breath Volume		1.925	01:46
Air Blank		0.000	01:46
Air Blank		0.000	01:48
Subject Sample	# 2	0.182	01:49
Breath Volume		1.722	01:49
Air Blank		0.000	01:50
Control Test		0.080	01:50
Air Blank		0.000	01:50
Diagnostics Check		OK	01:50

Data Download Date/Time: 08/10/2017 21:53

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLAIRE,J,L
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MOTTE,J
ARRESTING OFFICER AGENCY: Gulfport PD
LAST AGENCY INSPECTION: 06/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 17:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:23

SUBJECT'S LAST NAME: LAFAYE

DATE/TIME OF TEST: 07/02/2017 18:50:55

Diagnostics Check		OK	18:43
Air Blank		0.000	18:44
Control Test		0.079	18:44
Air Blank		0.000	18:45
Subject Sample	# 1	0.267	18:45
Breath Volume		2.085	18:45
Air Blank		0.000	18:46
Air Blank		0.000	18:48
Subject Sample	# 2	0.269	18:49
Breath Volume		2.101	18:49
Air Blank		0.000	18:50
Control Test		0.079	18:50
Air Blank		0.000	18:50
Diagnostics Check		OK	18:50

Data Download Date/Time: 07/28/2017 20:01

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLAIRE,J,L
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: FIELDEN,E
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 02:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: LUKE

DATE/TIME OF TEST: 07/03/2017 03:44:38

Diagnostics Check		OK	03:33	
Air Blank		0.000	03:33	
Control Test		0.079	03:34	
Air Blank		0.000	03:34	
Subject Sample	# 1		03:37	No Sample Provided
Breath Volume		0.000	03:37	
Air Blank		0.000	03:38	
Air Blank		0.000	03:40	
Subject Sample	# 2		03:43	No Sample Provided
Breath Volume		0.000	03:43	
Air Blank		0.000	03:43	
Control Test		0.079	03:44	
Air Blank		0.000	03:44	
Diagnostics Check		OK	03:44	

Data Download Date/Time: 07/28/2017 20:01

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLAIRE,J,L
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: FIELDEN,E
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 02:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: LUKE

DATE/TIME OF TEST: 07/03/2017 03:53:25

Diagnostics Check		OK	03:50	
Air Blank		0.000	03:50	
Control Test		0.079	03:51	
Air Blank		0.000	03:51	
Subject Sample	# 1		03:52	Subject Test Refused
Breath Volume		0.000	03:52	
Air Blank		0.000	03:52	
Control Test		0.079	03:52	
Air Blank		0.000	03:53	
Diagnostics Check		OK	03:53	

Data Download Date/Time: 07/28/2017 20:01

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILHELM,JAMES,D
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WILHELM,JAMES
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 03:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:18

SUBJECT'S LAST NAME: HARPER

DATE/TIME OF TEST: 07/04/2017 04:40:41

Diagnostics Check	OK	04:39	
Air Blank	0.000	04:39	
Control Test	0.079	04:39	
Air Blank		04:40	RFI Detect
Air Blank	0.000	04:40	

Data Download Date/Time: 07/28/2017 20:01

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILHELM,JAMES,D
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WILHELM,JAMES
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 03:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:18

SUBJECT'S LAST NAME: HARPER

DATE/TIME OF TEST: 07/04/2017 04:48:50

Diagnostics Check		OK	04:42
Air Blank		0.000	04:42
Control Test		0.079	04:42
Air Blank		0.000	04:43
Subject Sample	# 1	0.167	04:44
Breath Volume		2.023	04:44
Air Blank		0.000	04:44
Air Blank		0.000	04:46
Subject Sample	# 2	0.160	04:47
Breath Volume		2.167	04:47
Air Blank		0.000	04:47
Control Test		0.079	04:48
Air Blank		0.000	04:48
Diagnostics Check		OK	04:48

Data Download Date/Time: 07/28/2017 20:01

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLAIRE,J,L
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LEFFINGWELL,B
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 04:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:24

SUBJECT'S LAST NAME: VOYTOVYCH

DATE/TIME OF TEST: 07/17/2017 04:53:16

Diagnostics Check		OK	04:44	
Air Blank		0.000	04:45	
Control Test		0.080	04:45	
Air Blank		0.000	04:46	
Subject Sample	# 1		04:49	No Sample Provided
Breath Volume		0.000	04:49	
Air Blank		0.000	04:49	
Air Blank		0.000	04:51	
Subject Sample	# 2		04:51	Subject Test Refused
Breath Volume		0.000	04:51	
Air Blank		0.000	04:52	
Control Test		0.080	04:52	
Air Blank		0.000	04:53	
Diagnostics Check		OK	04:53	

Data Download Date/Time: 07/28/2017 20:01

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WEDE,B
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 03:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:44

SUBJECT'S LAST NAME: HABIB

DATE/TIME OF TEST: 07/22/2017 04:08:33

Diagnostics Check		OK	04:05	
Air Blank		0.000	04:06	
Control Test		0.079	04:06	
Air Blank		0.000	04:07	
Subject Sample	# 1		04:07	Subject Test Refused
Breath Volume		0.000	04:07	
Air Blank		0.000	04:07	
Control Test		0.079	04:07	
Air Blank		0.000	04:08	
Diagnostics Check		OK	04:08	

Data Download Date/Time: 07/28/2017 20:01

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001006

INSTRUMENT REGISTERED TO: INDIALANTIC P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIBSON,ZACHARY,A
BREATH TEST OPERATOR AGENCY: Indialantic PD
ARRESTING OFFICER: DULIN,HERBERT
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 06/07/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A1
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: HERNANDEZ MARTINEZ

DATE/TIME OF TEST: 07/07/2017 03:34:37

Diagnostics Check		OK	03:29	
Air Blank		0.000	03:29	
Control Test		0.078	03:30	
Air Blank		0.000	03:30	
Subject Sample	# 1		03:33	Subject Test Refused
Breath Volume		0.000	03:33	
Air Blank		0.000	03:33	
Control Test		0.079	03:34	
Air Blank		0.000	03:34	
Diagnostics Check		OK	03:34	

Data Download Date/Time: 07/31/2017 02:43

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001006

INSTRUMENT REGISTERED TO: INDIALANTIC P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLSTINE,SCOTT,
BREATH TEST OPERATOR AGENCY: Indialantic PD
ARRESTING OFFICER: EICHNER,AMANDA
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 06/07/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A1
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 08:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:34

SUBJECT'S LAST NAME: VANTERPOOL

DATE/TIME OF TEST: 07/15/2017 09:24:59

Diagnostics Check		OK	09:18
Air Blank		0.000	09:18
Control Test		0.079	09:18
Air Blank		0.000	09:19
Subject Sample	# 1	0.212	09:20
Breath Volume		2.687	09:20
Air Blank		0.000	09:21
Air Blank		0.000	09:22
Subject Sample	# 2	0.202	09:23
Breath Volume		2.527	09:23
Air Blank		0.000	09:24
Control Test		0.077	09:24
Air Blank		0.000	09:24
Diagnostics Check		OK	09:24

Data Download Date/Time: 07/31/2017 02:43

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001006

INSTRUMENT REGISTERED TO: INDIALANTIC P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VERDEROSA,TIFFANY,M
BREATH TEST OPERATOR AGENCY: Indialantic PD
ARRESTING OFFICER: VERDEROSA,TIFFANY
ARRESTING OFFICER AGENCY: Indialantic PD
LAST AGENCY INSPECTION: 06/07/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A1
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 10:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:45

SUBJECT'S LAST NAME: ROURKE

DATE/TIME OF TEST: 07/24/2017 11:44:52

Diagnostics Check		OK	11:37
Air Blank		0.000	11:38
Control Test		0.078	11:38
Air Blank		0.000	11:38
Subject Sample	# 1	0.000	11:40
Breath Volume		2.113	11:40
Air Blank		0.000	11:41
Air Blank		0.000	11:43
Subject Sample	# 2	0.000	11:43
Breath Volume		1.929	11:43
Air Blank		0.000	11:43
Control Test		0.077	11:44
Air Blank		0.000	11:44
Diagnostics Check		OK	11:44

Data Download Date/Time: 07/31/2017 02:43

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARREN,BILLY,J
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: WARREN,BILLY
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 06/16/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 684131
DRY GAS EXPIRY DATE: 01/23/2020
VIOLATION CODE: DUI
ARREST TIME: 00:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:08

SUBJECT'S LAST NAME: WHEELER

DATE/TIME OF TEST: 07/03/2017 01:46:10

Diagnostics Check		OK	01:39
Air Blank		0.000	01:39
Control Test		0.080	01:39
Air Blank		0.000	01:40
Subject Sample	# 1	0.196	01:41
Breath Volume		2.289	01:41
Air Blank		0.000	01:42
Air Blank		0.000	01:44
Subject Sample	# 2	0.203	01:44
Breath Volume		3.175	01:44
Air Blank		0.000	01:45
Control Test		0.083	01:45
Air Blank		0.000	01:46
Diagnostics Check		OK	01:46

Data Download Date/Time: 07/16/2017 09:09

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TUCKER,WILLIAM,
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: KALLMEYER,DANIEL
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 06/16/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 684131
DRY GAS EXPIRY DATE: 01/23/2020
VIOLATION CODE: DUI
ARREST TIME: 23:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: GLENN

DATE/TIME OF TEST: 07/09/2017 00:26:24

Diagnostics Check		OK	00:20
Air Blank		0.000	00:20
Control Test		0.082	00:21
Air Blank		0.000	00:21
Subject Sample	# 1	0.207	00:21
Breath Volume		2.214	00:21
Air Blank		0.000	00:22
Air Blank		0.000	00:24
Subject Sample	# 2	0.199	00:24
Breath Volume		2.464	00:24
Air Blank		0.000	00:25
Control Test		0.083	00:25
Air Blank		0.000	00:26
Diagnostics Check		OK	00:26

Data Download Date/Time: 07/16/2017 09:09

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ABU SOUD,AHMAD,A
BREATH TEST OPERATOR AGENCY: West Melbourne PD
ARRESTING OFFICER: ABU SOUD,AHMAD
ARRESTING OFFICER AGENCY: West Melbourne PD
LAST AGENCY INSPECTION: 06/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04216080A3
DRY GAS EXPIRY DATE: 03/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:47

SUBJECT'S LAST NAME: ST PIERRE

DATE/TIME OF TEST: 07/02/2017 22:16:43

Diagnostics Check		OK	22:10
Air Blank		0.000	22:11
Control Test		0.080	22:11
Air Blank		0.000	22:12
Subject Sample	# 1	0.000	22:12
Breath Volume		1.667	22:12
Air Blank		0.000	22:13
Air Blank		0.000	22:15
Subject Sample	# 2	0.000	22:15
Breath Volume		1.992	22:15
Air Blank		0.000	22:15
Control Test		0.079	22:16
Air Blank		0.000	22:16
Diagnostics Check		OK	22:16

Data Download Date/Time: 07/31/2017 08:41

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ABU SOUD,AHMAD,A
BREATH TEST OPERATOR AGENCY: West Melbourne PD
ARRESTING OFFICER: THOMAS,ADREAN
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 06/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04216080A3
DRY GAS EXPIRY DATE: 03/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:22

SUBJECT'S LAST NAME: LUNA

DATE/TIME OF TEST: 07/05/2017 20:53:41

Diagnostics Check		OK	20:46
Air Blank		0.000	20:46
Control Test		0.080	20:47
Air Blank		0.000	20:47
Subject Sample	# 1	0.163	20:48
Breath Volume		1.710	20:48
Air Blank		0.000	20:49
Air Blank		0.000	20:51
Subject Sample	# 2	0.170	20:52
Breath Volume		1.757	20:52
Air Blank		0.000	20:52
Control Test		0.078	20:53
Air Blank		0.000	20:53
Diagnostics Check		OK	20:53

Data Download Date/Time: 07/31/2017 08:41

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ABU SOUD,AHMAD,A
BREATH TEST OPERATOR AGENCY: West Melbourne PD
ARRESTING OFFICER: RUTHERFORD,CHRIS
ARRESTING OFFICER AGENCY: West Melbourne PD
LAST AGENCY INSPECTION: 06/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04216080A3
DRY GAS EXPIRY DATE: 03/05/2018
VIOLATION CODE: DUI
ARREST TIME: 18:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:05

SUBJECT'S LAST NAME: LEE

DATE/TIME OF TEST: 07/11/2017 19:33:41

Diagnostics Check		OK	19:27
Air Blank		0.000	19:27
Control Test		0.079	19:28
Air Blank		0.000	19:28
Subject Sample	# 1	0.275	19:29
Breath Volume		3.503	19:29
Air Blank		0.000	19:30
Air Blank		0.000	19:31
Subject Sample	# 2	0.288	19:32
Breath Volume		3.503	19:32
Air Blank		0.000	19:32
Control Test		0.077	19:33
Air Blank		0.000	19:33
Diagnostics Check		OK	19:33

Data Download Date/Time: 07/31/2017 08:41

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUTTEBO,KENT,N
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: GUTTEBO,KENT
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 06/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04216080A3
DRY GAS EXPIRY DATE: 03/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:30

SUBJECT'S LAST NAME: IHNKEN

DATE/TIME OF TEST: 07/12/2017 21:03:39

Diagnostics Check		OK	20:57
Air Blank		0.000	20:58
Control Test		0.079	20:58
Air Blank		0.000	20:58
Subject Sample	# 1	0.165	20:59
Breath Volume		2.644	20:59
Air Blank		0.000	20:59
Air Blank		0.000	21:01
Subject Sample	# 2	0.164	21:02
Breath Volume		2.730	21:02
Air Blank		0.000	21:02
Control Test		0.079	21:03
Air Blank		0.000	21:03
Diagnostics Check		OK	21:03

Data Download Date/Time: 07/31/2017 08:41

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OBYRNE,JAMES,W
BREATH TEST OPERATOR AGENCY: West Melbourne PD
ARRESTING OFFICER: OBYRE,JAMES
ARRESTING OFFICER AGENCY: West Melbourne PD
LAST AGENCY INSPECTION: 06/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04216080A3
DRY GAS EXPIRY DATE: 03/05/2018
VIOLATION CODE: DUI
ARREST TIME: 14:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:45

SUBJECT'S LAST NAME: HARTSOCK

DATE/TIME OF TEST: 07/14/2017 15:14:50

Diagnostics Check		OK	15:08
Air Blank		0.000	15:09
Control Test		0.079	15:09
Air Blank		0.000	15:09
Subject Sample	# 1	0.240	15:10
Breath Volume		3.656	15:10
Air Blank		0.000	15:11
Air Blank		0.000	15:12
Subject Sample	# 2	0.238	15:13
Breath Volume		3.480	15:13
Air Blank		0.000	15:13
Control Test		0.078	15:14
Air Blank		0.000	15:14
Diagnostics Check		OK	15:14

Data Download Date/Time: 07/31/2017 08:41

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ABU SOUD,A,A
BREATH TEST OPERATOR AGENCY: West Melbourne PD
ARRESTING OFFICER: RYBA,L
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 06/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04216080A3
DRY GAS EXPIRY DATE: 03/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:19

SUBJECT'S LAST NAME: CARROLL

DATE/TIME OF TEST: 07/17/2017 00:46:19

Diagnostics Check		OK	00:40
Air Blank		0.000	00:40
Control Test		0.079	00:41
Air Blank		0.000	00:41
Subject Sample	# 1	0.000	00:42
Breath Volume		1.687	00:42
Air Blank		0.000	00:42
Air Blank		0.000	00:44
Subject Sample	# 2	0.000	00:44
Breath Volume		1.285	00:44
Air Blank		0.000	00:45
Control Test		0.079	00:45
Air Blank		0.000	00:46
Diagnostics Check		OK	00:46

Data Download Date/Time: 07/31/2017 08:41

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ABU SOUD,AHMAD,A
BREATH TEST OPERATOR AGENCY: West Melbourne PD
ARRESTING OFFICER: RUTHERFORD,CHRIS
ARRESTING OFFICER AGENCY: West Melbourne PD
LAST AGENCY INSPECTION: 06/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04216080A3
DRY GAS EXPIRY DATE: 03/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:56

SUBJECT'S LAST NAME: CRANE

DATE/TIME OF TEST: 07/17/2017 02:23:15

Diagnostics Check		OK	02:16
Air Blank		0.000	02:17
Control Test		0.078	02:17
Air Blank		0.000	02:18
Subject Sample	# 1	0.182	02:18
Breath Volume		1.960	02:18
Air Blank		0.000	02:19
Air Blank		0.000	02:21
Subject Sample	# 2	0.184	02:21
Breath Volume		1.957	02:21
Air Blank		0.000	02:22
Control Test		0.077	02:22
Air Blank		0.000	02:23
Diagnostics Check		OK	02:23

Data Download Date/Time: 07/31/2017 08:41

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ABU SOUD,AHMAD,A
BREATH TEST OPERATOR AGENCY: West Melbourne PD
ARRESTING OFFICER: RUTHERFORD,CHRIS
ARRESTING OFFICER AGENCY: West Melbourne PD
LAST AGENCY INSPECTION: 06/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04216080A3
DRY GAS EXPIRY DATE: 03/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:53

SUBJECT'S LAST NAME: PERNAL

DATE/TIME OF TEST: 07/26/2017 03:25:54

Diagnostics Check		OK	03:14	
Air Blank		0.000	03:14	
Control Test		0.079	03:15	
Air Blank		0.000	03:15	
Subject Sample	# 1		03:18	No Sample Provided
Breath Volume		0.000	03:18	
Air Blank		0.000	03:19	
Air Blank		0.000	03:21	
Subject Sample	# 2		03:24	No Sample Provided
Breath Volume		0.000	03:24	
Air Blank		0.000	03:25	
Control Test		0.079	03:25	
Air Blank		0.000	03:25	
Diagnostics Check		OK	03:25	

Data Download Date/Time: 07/31/2017 08:41

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001009

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMILTON,J,W
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: HAMILTON,J
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 06/23/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A2
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:48

SUBJECT'S LAST NAME: PLATANO

DATE/TIME OF TEST: 07/09/2017 00:27:20

Diagnostics Check		OK	00:20
Air Blank		0.000	00:21
Control Test		0.079	00:21
Air Blank		0.000	00:22
Subject Sample	# 1	0.000	00:22
Breath Volume		2.511	00:22
Air Blank		0.000	00:23
Air Blank		0.000	00:25
Subject Sample	# 2	0.000	00:25
Breath Volume		2.218	00:25
Air Blank		0.000	00:26
Control Test		0.079	00:26
Air Blank		0.000	00:27
Diagnostics Check		OK	00:27

Data Download Date/Time: 07/12/2017 14:50

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAHIM,A,
BREATH TEST OPERATOR AGENCY: Mayport Naval Base
ARRESTING OFFICER: HOLLAND,D
ARRESTING OFFICER AGENCY: Mayport Naval Base
LAST AGENCY INSPECTION: 06/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:40

SUBJECT'S LAST NAME: HILLER

DATE/TIME OF TEST: 07/03/2017 21:19:15

Diagnostics Check		OK	21:13	
Air Blank		0.000	21:14	
Control Test		0.077	21:14	
Air Blank		0.000	21:14	
Subject Sample	# 1	0.173	21:18	Volume Not Met
Breath Volume		1.046	21:18	
Air Blank			21:18	Purge Fail
Air Blank		0.000	21:19	

Data Download Date/Time: 07/27/2017 14:51

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOOS,JAMES,P
BREATH TEST OPERATOR AGENCY: Mayport Naval Base
ARRESTING OFFICER: MARRERO,JOSE
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:15

SUBJECT'S LAST NAME: OFFUTT

DATE/TIME OF TEST: 07/10/2017 23:01:40

Diagnostics Check		OK	22:55
Air Blank		0.000	22:55
Control Test		0.079	22:55
Air Blank		0.000	22:56
Subject Sample	# 1	0.105	22:57
Breath Volume		2.691	22:57
Air Blank		0.000	22:57
Air Blank		0.000	22:59
Subject Sample	# 2	0.102	22:59
Breath Volume		2.605	22:59
Air Blank		0.000	23:00
Control Test		0.077	23:01
Air Blank		0.000	23:01
Diagnostics Check		OK	23:01

Data Download Date/Time: 07/27/2017 14:51

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OCASIO,JOSE,A
BREATH TEST OPERATOR AGENCY: Mayport Naval Base
ARRESTING OFFICER: MCKINNEY,RHIANNON
ARRESTING OFFICER AGENCY: Mayport Naval Base
LAST AGENCY INSPECTION: 06/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: FRIESENHAHN

DATE/TIME OF TEST: 07/16/2017 00:43:14

Diagnostics Check		OK	00:36
Air Blank		0.000	00:36
Control Test		0.079	00:37
Air Blank		0.000	00:37
Subject Sample	# 1	0.072	00:38
Breath Volume		1.609	00:38
Air Blank		0.000	00:38
Air Blank		0.000	00:40
Subject Sample	# 2	0.071	00:41
Breath Volume		2.730	00:41
Air Blank		0.000	00:42
Control Test		0.079	00:42
Air Blank		0.000	00:43
Diagnostics Check		OK	00:43

Data Download Date/Time: 07/27/2017 14:51

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JASIUKIEWICZ,MICHAEL,P
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: JASIUKIEWICZ,MICHAEL
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 06/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:53

SUBJECT'S LAST NAME: CONCHA CASTANEDA

DATE/TIME OF TEST: 07/01/2017 01:21:22

Diagnostics Check		OK	01:14
Air Blank		0.000	01:14
Control Test		0.079	01:15
Air Blank		0.000	01:15
Subject Sample	# 1	0.284	01:16
Breath Volume		3.464	01:16
Air Blank		0.000	01:17
Air Blank		0.000	01:19
Subject Sample	# 2	0.276	01:19
Breath Volume		3.699	01:19
Air Blank		0.000	01:20
Control Test		0.080	01:20
Air Blank		0.000	01:21
Diagnostics Check		OK	01:21

Data Download Date/Time: 07/14/2017 10:32

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROLLESTON,JACOB,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: ROLLESTON,JACOB
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 06/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: MORGAN

DATE/TIME OF TEST: 07/01/2017 02:22:04

Diagnostics Check		OK	02:19	
Air Blank		0.000	02:19	
Control Test		0.080	02:19	
Air Blank		0.000	02:20	
Subject Sample	# 1		02:20	Subject Test Refused
Breath Volume		0.000	02:20	
Air Blank		0.000	02:21	
Control Test		0.079	02:21	
Air Blank		0.000	02:21	
Diagnostics Check		OK	02:22	

Data Download Date/Time: 07/14/2017 10:32

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHEA,CPL,J
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: SHEA,JOHN
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 06/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:18

SUBJECT'S LAST NAME: HARPER

DATE/TIME OF TEST: 07/01/2017 02:46:57

Diagnostics Check		OK	02:40
Air Blank		0.000	02:40
Control Test		0.079	02:41
Air Blank		0.000	02:41
Subject Sample	# 1	0.169	02:42
Breath Volume		4.656	02:42
Air Blank		0.000	02:43
Air Blank		0.000	02:45
Subject Sample	# 2	0.173	02:45
Breath Volume		4.503	02:45
Air Blank		0.000	02:46
Control Test		0.081	02:46
Air Blank		0.000	02:46
Diagnostics Check		OK	02:46

Data Download Date/Time: 07/14/2017 10:32

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JASIUKIEWICZ,MICHAEL,P
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: JASIUKIEWICZ,MICHAEL
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 06/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: ALLEN-MILLS

DATE/TIME OF TEST: 07/02/2017 03:34:50

Diagnostics Check		OK	03:28
Air Blank		0.000	03:29
Control Test		0.079	03:29
Air Blank		0.000	03:29
Subject Sample	# 1	0.194	03:30
Breath Volume		2.660	03:30
Air Blank		0.000	03:31
Air Blank		0.000	03:32
Subject Sample	# 2	0.199	03:33
Breath Volume		2.648	03:33
Air Blank		0.000	03:33
Control Test		0.081	03:34
Air Blank		0.000	03:34
Diagnostics Check		OK	03:34

Data Download Date/Time: 07/14/2017 10:32

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHMIDT,RICHARD,W
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: SMITH,SEVWANE
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 06/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:21

SUBJECT'S LAST NAME: PARRA

DATE/TIME OF TEST: 07/03/2017 05:46:07

Diagnostics Check		OK	05:41	
Air Blank		0.000	05:42	
Control Test		0.079	05:42	
Air Blank		0.000	05:42	
Subject Sample	# 1		05:44	Subject Test Refused
Breath Volume		0.000	05:44	
Air Blank		0.000	05:45	
Control Test		0.078	05:45	
Air Blank		0.000	05:46	
Diagnostics Check		OK	05:46	

Data Download Date/Time: 07/14/2017 10:32

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHEA,CPL,J
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: SHEA,JOHN
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 06/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:07

SUBJECT'S LAST NAME: BREINGAN

DATE/TIME OF TEST: 07/04/2017 03:30:40

Diagnostics Check		OK	03:27	
Air Blank		0.000	03:28	
Control Test		0.079	03:28	
Air Blank		0.000	03:29	
Subject Sample	# 1		03:29	Subject Test Refused
Breath Volume		0.000	03:29	
Air Blank		0.000	03:29	
Control Test		0.079	03:30	
Air Blank		0.000	03:30	
Diagnostics Check		OK	03:30	

Data Download Date/Time: 07/14/2017 10:32

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHEA,CPL,J
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: SHEA,JOHN
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 06/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:38

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 07/04/2017 22:05:36

Diagnostics Check		OK	21:59
Air Blank		0.000	21:59
Control Test		0.079	22:00
Air Blank		0.000	22:00
Subject Sample	# 1	0.000	22:01
Breath Volume		1.308	22:01
Air Blank		0.000	22:01
Air Blank		0.000	22:03
Subject Sample	# 2	0.000	22:04
Breath Volume		1.722	22:04
Air Blank		0.000	22:04
Control Test		0.082	22:05
Air Blank		0.000	22:05
Diagnostics Check		OK	22:05

Data Download Date/Time: 07/14/2017 10:32

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JASIUKIEWICZ,MICHAEL,P
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: JASIUKIEWICZ,MICHAEL
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 06/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:18

SUBJECT'S LAST NAME: HUBBARD

DATE/TIME OF TEST: 07/09/2017 00:40:55

Diagnostics Check		OK	00:38	
Air Blank		0.000	00:38	
Control Test		0.078	00:38	
Air Blank		0.000	00:39	
Subject Sample	# 1		00:39	Subject Test Refused
Breath Volume		0.000	00:39	
Air Blank		0.000	00:40	
Control Test		0.079	00:40	
Air Blank		0.000	00:40	
Diagnostics Check		OK	00:40	

Data Download Date/Time: 07/14/2017 10:32

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOYLE,THOMAS,E
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: BOYLE,THOMAS
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 06/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:07

SUBJECT'S LAST NAME: USATEGUI II

DATE/TIME OF TEST: 07/09/2017 21:33:32

Diagnostics Check		OK	21:27
Air Blank		0.000	21:27
Control Test		0.079	21:28
Air Blank		0.000	21:28
Subject Sample	# 1	0.172	21:29
Breath Volume		1.148	21:29
Air Blank		0.000	21:29
Air Blank		0.000	21:31
Subject Sample	# 2	0.168	21:32
Breath Volume		1.492	21:32
Air Blank		0.000	21:32
Control Test		0.080	21:32
Air Blank		0.000	21:33
Diagnostics Check		OK	21:33

Data Download Date/Time: 07/14/2017 10:32

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHEA,CPL,J
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: SHEA,JOHN
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 06/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: HOLCOMB

DATE/TIME OF TEST: 07/12/2017 23:13:53

Diagnostics Check		OK	23:11	
Air Blank		0.000	23:11	
Control Test		0.079	23:11	
Air Blank		0.000	23:12	
Subject Sample	# 1		23:12	Subject Test Refused
Breath Volume		0.000	23:12	
Air Blank		0.000	23:12	
Control Test		0.079	23:13	
Air Blank		0.000	23:13	
Diagnostics Check		OK	23:13	

Data Download Date/Time: 07/14/2017 10:32

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHEA,CPL,J
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: SHEA,JOHN
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 06/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:24

SUBJECT'S LAST NAME: DELONG

DATE/TIME OF TEST: 07/13/2017 03:50:26

Diagnostics Check		OK	03:44
Air Blank		0.000	03:44
Control Test		0.079	03:45
Air Blank		0.000	03:45
Subject Sample	# 1	0.037	03:46
Breath Volume		2.781	03:46
Air Blank		0.000	03:46
Air Blank		0.000	03:48
Subject Sample	# 2	0.033	03:49
Breath Volume		1.109	03:49
Air Blank		0.000	03:49
Control Test		0.080	03:49
Air Blank		0.000	03:50
Diagnostics Check		OK	03:50

Data Download Date/Time: 07/14/2017 10:32

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JASIUKIEWICZ,MICHAEL,P
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: PERRAULT,ROBERT
ARRESTING OFFICER AGENCY: Kenneth City PD
LAST AGENCY INSPECTION: 06/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:42

SUBJECT'S LAST NAME: BERUBE

DATE/TIME OF TEST: 07/13/2017 23:08:50

Diagnostics Check		OK	23:02
Air Blank		0.000	23:03
Control Test		0.079	23:03
Air Blank		0.000	23:03
Subject Sample	# 1	0.210	23:04
Breath Volume		3.140	23:04
Air Blank		0.000	23:04
Air Blank		0.000	23:06
Subject Sample	# 2	0.217	23:07
Breath Volume		2.707	23:07
Air Blank		0.000	23:07
Control Test		0.079	23:08
Air Blank		0.000	23:08
Diagnostics Check		OK	23:08

Data Download Date/Time: 07/14/2017 10:32

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROLLESTON,JACOB,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: ROLLESTON,JACOB
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 07/14/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:49

SUBJECT'S LAST NAME: HRAMIKA

DATE/TIME OF TEST: 07/15/2017 00:12:19

Diagnostics Check		OK	00:09	
Air Blank		0.000	00:09	
Control Test		0.079	00:10	
Air Blank		0.000	00:10	
Subject Sample	# 1		00:10	Subject Test Refused
Breath Volume		0.000	00:10	
Air Blank		0.000	00:11	
Control Test		0.079	00:11	
Air Blank		0.000	00:12	
Diagnostics Check		OK	00:12	

Data Download Date/Time: 08/09/2017 10:12

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SORICE,DOMINIC,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: BOATWRIGHT,JAMARIOUS
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 07/14/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 16:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:17

SUBJECT'S LAST NAME: HASELTINE

DATE/TIME OF TEST: 07/19/2017 17:45:23

Diagnostics Check		OK	17:39
Air Blank		0.000	17:39
Control Test		0.078	17:39
Air Blank		0.000	17:40
Subject Sample	# 1	0.209	17:40
Breath Volume		2.777	17:40
Air Blank		0.000	17:41
Air Blank		0.000	17:43
Subject Sample	# 2	0.208	17:43
Breath Volume		2.742	17:43
Air Blank		0.000	17:44
Control Test		0.079	17:44
Air Blank		0.000	17:45
Diagnostics Check		OK	17:45

Data Download Date/Time: 08/09/2017 10:12

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROLLESTON,JACOB,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: ROLLESTON,JACOB
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 07/14/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:42

SUBJECT'S LAST NAME: ROBERSON

DATE/TIME OF TEST: 07/20/2017 22:05:03

Diagnostics Check		OK	22:02	
Air Blank		0.000	22:02	
Control Test		0.079	22:03	
Air Blank		0.000	22:03	
Subject Sample	# 1		22:03	Subject Test Refused
Breath Volume		0.000	22:03	
Air Blank		0.000	22:04	
Control Test		0.078	22:04	
Air Blank		0.000	22:04	
Diagnostics Check		OK	22:05	

Data Download Date/Time: 08/09/2017 10:12

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JASIUKIEWICZ,MICHAEL,P
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: JASIUKIEWICZ,MICHAEL
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 07/14/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:52

SUBJECT'S LAST NAME: ODDO

DATE/TIME OF TEST: 07/21/2017 03:15:12

Diagnostics Check		OK	03:12	
Air Blank		0.000	03:12	
Control Test		0.079	03:13	
Air Blank		0.000	03:13	
Subject Sample	# 1		03:13	Subject Test Refused
Breath Volume		0.000	03:13	
Air Blank		0.000	03:14	
Control Test		0.078	03:14	
Air Blank		0.000	03:15	
Diagnostics Check		OK	03:15	

Data Download Date/Time: 08/09/2017 10:12

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JASIUKIEWICZ,MICHAEL,P
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: JASIUKIEWICZ,MICHAEL
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 07/14/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 14:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:02

SUBJECT'S LAST NAME: MAKSYMOWSKA

DATE/TIME OF TEST: 07/23/2017 15:28:52

Diagnostics Check		OK	15:22
Air Blank		0.000	15:23
Control Test		0.079	15:23
Air Blank		0.000	15:24
Subject Sample	# 1	0.000	15:24
Breath Volume		3.183	15:24
Air Blank		0.000	15:25
Air Blank		0.000	15:27
Subject Sample	# 2	0.000	15:27
Breath Volume		3.300	15:27
Air Blank		0.000	15:27
Control Test		0.081	15:28
Air Blank		0.000	15:28
Diagnostics Check		OK	15:28

Data Download Date/Time: 08/09/2017 10:12

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHEA,CPL,J
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: SHEA,JOHN
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 07/14/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:21

SUBJECT'S LAST NAME: STOVER

DATE/TIME OF TEST: 07/25/2017 02:45:33

Diagnostics Check		OK	02:42	
Air Blank		0.000	02:43	
Control Test		0.079	02:43	
Air Blank		0.000	02:43	
Subject Sample	# 1		02:44	Subject Test Refused
Breath Volume		0.000	02:44	
Air Blank		0.000	02:44	
Control Test		0.078	02:44	
Air Blank		0.000	02:45	
Diagnostics Check		OK	02:45	

Data Download Date/Time: 08/09/2017 10:12

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROLLESTON,JACOB,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: ROLLESTON,JACOB
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 07/14/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 18:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:56

SUBJECT'S LAST NAME: ALLEN

DATE/TIME OF TEST: 07/26/2017 19:27:59

Diagnostics Check		OK	19:21	
Air Blank		0.000	19:21	
Control Test		0.079	19:22	
Air Blank		0.000	19:22	
Subject Sample	# 1	0.362	19:23	
Breath Volume		2.218	19:23	
Air Blank		0.000	19:23	
Air Blank		0.000	19:25	
Subject Sample	# 2	0.373	19:26	
Breath Volume		1.632	19:26	
Air Blank		0.000	19:26	
Control Test		0.079	19:26	
Air Blank		0.000	19:27	
Diagnostics Check		Fail	19:27	Temperature Regulation Test
Air Blank		0.000	19:27	

Data Download Date/Time: 08/09/2017 10:12

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROLLESTON,JACOB,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: ROLLESTON,JACOB
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 07/14/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 18:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:56

SUBJECT'S LAST NAME: ALLEN

DATE/TIME OF TEST: 07/26/2017 19:36:03

Diagnostics Check		OK	19:30
Air Blank		0.000	19:30
Control Test		0.079	19:30
Air Blank		0.000	19:31
Subject Sample	# 1	0.373	19:31
Breath Volume		2.222	19:31
Air Blank		0.000	19:32
Air Blank		0.000	19:34
Subject Sample	# 2	0.362	19:34
Breath Volume		2.265	19:34
Air Blank		0.000	19:35
Control Test		0.078	19:35
Air Blank		0.000	19:35
Diagnostics Check		OK	19:36

Data Download Date/Time: 08/09/2017 10:12

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROLLESTON,JACOB,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: ROLLESTON,JACOB
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 07/14/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: JOTA

DATE/TIME OF TEST: 07/27/2017 01:34:00

Diagnostics Check		OK	01:30	
Air Blank		0.000	01:31	
Control Test		0.078	01:31	
Air Blank		0.000	01:31	
Subject Sample	# 1		01:32	Subject Test Refused
Breath Volume		0.000	01:32	
Air Blank		0.000	01:33	
Control Test		0.079	01:33	
Air Blank		0.000	01:33	
Diagnostics Check		OK	01:34	

Data Download Date/Time: 08/09/2017 10:12

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001043

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROE,ROBERT,D
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: RHODEN,DAVID
ARRESTING OFFICER AGENCY: Lake Placid PD
LAST AGENCY INSPECTION: 06/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: PROB/PAROL
ARREST TIME: 01:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: STEPHENSON

DATE/TIME OF TEST: 07/19/2017 02:01:26

Diagnostics Check		OK	01:54
Air Blank		0.000	01:55
Control Test		0.078	01:55
Air Blank		0.000	01:56
Subject Sample	# 1	0.025	01:56
Breath Volume		1.296	01:56
Air Blank		0.000	01:57
Air Blank		0.000	01:59
Subject Sample	# 2	0.025	02:00
Breath Volume		1.226	02:00
Air Blank		0.000	02:00
Control Test		0.077	02:00
Air Blank		0.000	02:01
Diagnostics Check		OK	02:01

Data Download Date/Time: 07/24/2017 15:31

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001043

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRACKETT,MARK,O
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: BRACKETT,MARK
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 06/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: ZERO TOL
ARREST TIME: 10:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:45

SUBJECT'S LAST NAME: WINDLAND

DATE/TIME OF TEST: 07/19/2017 11:14:50

Diagnostics Check		OK	11:08
Air Blank		0.000	11:08
Control Test		0.077	11:09
Air Blank		0.000	11:09
Subject Sample	# 1	0.000	11:10
Breath Volume		2.039	11:10
Air Blank		0.000	11:10
Air Blank		0.000	11:12
Subject Sample	# 2	0.000	11:13
Breath Volume		2.175	11:13
Air Blank		0.000	11:13
Control Test		0.077	11:14
Air Blank		0.000	11:14
Diagnostics Check		OK	11:14

Data Download Date/Time: 07/24/2017 15:31

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001043

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRACKETT,MARK,O
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: BRACKETT,MARK
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 06/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: ZERO TOL
ARREST TIME: 11:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 11:35

SUBJECT'S LAST NAME: HECKLER

DATE/TIME OF TEST: 07/19/2017 12:04:57

Diagnostics Check		OK	11:58
Air Blank		0.000	11:58
Control Test		0.078	11:59
Air Blank		0.000	11:59
Subject Sample	# 1	0.000	12:00
Breath Volume		2.539	12:00
Air Blank		0.000	12:01
Air Blank		0.000	12:03
Subject Sample	# 2	0.000	12:03
Breath Volume		1.894	12:03
Air Blank		0.000	12:04
Control Test		0.077	12:04
Air Blank		0.000	12:04
Diagnostics Check		OK	12:04

Data Download Date/Time: 07/24/2017 15:31

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001043

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PECK,JOHN,R
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: PECK,JOHN
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 07/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: ZERO TOL
ARREST TIME: 16:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:57

SUBJECT'S LAST NAME: MCGEE

DATE/TIME OF TEST: 07/26/2017 17:28:21

Diagnostics Check		OK	17:22
Air Blank		0.000	17:22
Control Test		0.078	17:23
Air Blank		0.000	17:23
Subject Sample	# 1	0.000	17:24
Breath Volume		1.433	17:24
Air Blank		0.000	17:24
Air Blank		0.000	17:26
Subject Sample	# 2	0.000	17:27
Breath Volume		1.148	17:27
Air Blank		0.000	17:27
Control Test		0.076	17:27
Air Blank		0.000	17:28
Diagnostics Check		OK	17:28

Data Download Date/Time: 08/25/2017 12:30

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001043

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,WILLIAM,M
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: SALVADOR,DEPUTY
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 07/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 07/30/2017 05:01:00

Diagnostics Check		OK	04:57	
Air Blank		0.000	04:58	
Control Test		0.078	04:58	
Air Blank		0.000	04:59	
Subject Sample	# 1		04:59	Subject Test Refused
Breath Volume		0.000	04:59	
Air Blank		0.000	05:00	
Control Test		0.077	05:00	
Air Blank		0.000	05:00	
Diagnostics Check		OK	05:01	

Data Download Date/Time: 08/25/2017 12:30

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORALES,DAVID,J
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf
ARRESTING OFFICER: MORALES,DAVID
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf
LAST AGENCY INSPECTION: 06/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 01:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: KEITT

DATE/TIME OF TEST: 07/01/2017 02:44:31

Diagnostics Check		OK	02:41	
Air Blank		0.000	02:42	
Control Test		0.081	02:42	
Air Blank		0.000	02:43	
Subject Sample	# 1		02:43	Subject Test Refused
Breath Volume		0.000	02:43	
Air Blank		0.000	02:43	
Control Test		0.081	02:43	
Air Blank		0.000	02:44	
Diagnostics Check		OK	02:44	

Data Download Date/Time: 07/26/2017 12:22

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESCOBAR,ALEJANDRO,
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf
ARRESTING OFFICER: ESCOBAR,ALEJANDRO
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf
LAST AGENCY INSPECTION: 06/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 00:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:42

SUBJECT'S LAST NAME: ZAREWICZ

DATE/TIME OF TEST: 07/02/2017 01:18:10

Diagnostics Check		OK	01:12
Air Blank		0.000	01:12
Control Test		0.081	01:12
Air Blank		0.000	01:13
Subject Sample	# 1	0.257	01:13
Breath Volume		2.289	01:13
Air Blank		0.000	01:14
Air Blank		0.000	01:16
Subject Sample	# 2	0.256	01:16
Breath Volume		2.121	01:16
Air Blank		0.000	01:17
Control Test		0.081	01:17
Air Blank		0.000	01:18
Diagnostics Check		OK	01:18

Data Download Date/Time: 07/26/2017 12:22

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESCOBAR,ALEJANDRO,
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf
ARRESTING OFFICER: ESCOBAR,ALEJANDRO
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf
LAST AGENCY INSPECTION: 06/26/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:47

SUBJECT'S LAST NAME: PROCOPIO

DATE/TIME OF TEST: 07/17/2017 02:18:19

Diagnostics Check		OK	02:15	
Air Blank		0.000	02:16	
Control Test		0.081	02:16	
Air Blank		0.000	02:16	
Subject Sample	# 1		02:16	Subject Test Refused
Breath Volume		0.000	02:16	
Air Blank		0.000	02:17	
Control Test		0.080	02:17	
Air Blank		0.000	02:18	
Diagnostics Check		OK	02:18	

Data Download Date/Time: 07/26/2017 12:22

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PADRON,RICARDO,
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf
ARRESTING OFFICER: HENRY,GAVIN
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf
LAST AGENCY INSPECTION: 06/26/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 02:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: BODRICK

DATE/TIME OF TEST: 07/17/2017 03:28:29

Diagnostics Check		OK	03:20
Air Blank		0.000	03:21
Control Test		0.080	03:21
Air Blank		0.000	03:21
Subject Sample	# 1	0.211	03:23
Breath Volume		2.332	03:23
Air Blank		0.000	03:24
Air Blank		0.000	03:26
Subject Sample	# 2	0.197	03:26
Breath Volume		1.378	03:26
Air Blank		0.000	03:27
Control Test		0.079	03:27
Air Blank		0.000	03:28
Diagnostics Check		OK	03:28

Data Download Date/Time: 07/26/2017 12:22

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LARA,JENNIFER,A
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf
ARRESTING OFFICER: DAVIS,DONALD
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf
LAST AGENCY INSPECTION: 06/26/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 09:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:25

SUBJECT'S LAST NAME: PHIPPSCAMILO

DATE/TIME OF TEST: 07/18/2017 10:05:38

Diagnostics Check		OK	09:59
Air Blank		0.000	09:59
Control Test		0.081	10:00
Air Blank		0.000	10:00
Subject Sample	# 1	0.249	10:01
Breath Volume		1.652	10:01
Air Blank		0.000	10:01
Air Blank		0.000	10:03
Subject Sample	# 2	0.258	10:04
Breath Volume		1.468	10:04
Air Blank		0.000	10:04
Control Test		0.080	10:05
Air Blank		0.000	10:05
Diagnostics Check		OK	10:05

Data Download Date/Time: 07/26/2017 12:22

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESCOBAR,ALEJANDRO,
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf
ARRESTING OFFICER: ESCOBAR,ALEJANDRO
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf
LAST AGENCY INSPECTION: 06/26/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 00:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: GUERRIDO MONSERRATE

DATE/TIME OF TEST: 07/21/2017 01:28:49

Diagnostics Check		OK	01:22
Air Blank		0.000	01:22
Control Test		0.080	01:23
Air Blank		0.000	01:23
Subject Sample	# 1	0.229	01:24
Breath Volume		2.113	01:24
Air Blank		0.000	01:24
Air Blank		0.000	01:26
Subject Sample	# 2	0.234	01:27
Breath Volume		2.929	01:27
Air Blank		0.000	01:27
Control Test		0.079	01:28
Air Blank		0.000	01:28
Diagnostics Check		OK	01:28

Data Download Date/Time: 07/26/2017 12:22

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PADRON,RICARDO,
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf
ARRESTING OFFICER: HENRY,GAVIN
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf
LAST AGENCY INSPECTION: 06/26/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 02:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: LUDWIN

DATE/TIME OF TEST: 07/21/2017 03:02:21

Diagnostics Check		OK	02:56
Air Blank		0.000	02:56
Control Test		0.080	02:56
Air Blank		0.000	02:57
Subject Sample	# 1	0.217	02:57
Breath Volume		1.734	02:57
Air Blank		0.000	02:58
Air Blank		0.000	03:00
Subject Sample	# 2	0.214	03:00
Breath Volume		1.621	03:00
Air Blank		0.000	03:01
Control Test		0.080	03:01
Air Blank		0.000	03:02
Diagnostics Check		OK	03:02

Data Download Date/Time: 07/26/2017 12:22

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESCOBAR,ALEJANDRO,
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf
ARRESTING OFFICER: ESCOBAR,ALEJANDRO
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf
LAST AGENCY INSPECTION: 06/26/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 03:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: STOUT

DATE/TIME OF TEST: 07/21/2017 04:36:45

Diagnostics Check		OK	04:30
Air Blank		0.000	04:31
Control Test		0.080	04:31
Air Blank		0.000	04:32
Subject Sample	# 1	0.187	04:32
Breath Volume		3.121	04:32
Air Blank		0.000	04:32
Air Blank		0.000	04:34
Subject Sample	# 2	0.182	04:35
Breath Volume		2.933	04:35
Air Blank		0.000	04:35
Control Test		0.080	04:36
Air Blank		0.000	04:36
Diagnostics Check		OK	04:36

Data Download Date/Time: 07/26/2017 12:22

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESCOBAR,ALEJANDRO,
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf
ARRESTING OFFICER: ESCOBAR,ALEJANDRO
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf
LAST AGENCY INSPECTION: 06/26/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 20:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:58

SUBJECT'S LAST NAME: LEVY

DATE/TIME OF TEST: 07/21/2017 21:28:59

Diagnostics Check		OK	21:26	
Air Blank		0.000	21:26	
Control Test		0.080	21:26	
Air Blank		0.000	21:27	
Subject Sample	# 1		21:27	Subject Test Refused
Breath Volume		0.000	21:27	
Air Blank		0.000	21:28	
Control Test		0.081	21:28	
Air Blank		0.000	21:28	
Diagnostics Check		OK	21:28	

Data Download Date/Time: 07/26/2017 12:22

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESCOBAR,ALEJANDRO,
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf
ARRESTING OFFICER: ESCOBAR,ALEJANDRO
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf
LAST AGENCY INSPECTION: 06/26/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 03:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:28

SUBJECT'S LAST NAME: KING

DATE/TIME OF TEST: 07/21/2017 04:10:50

Diagnostics Check	OK	04:09	
Air Blank	0.000	04:09	
Control Test	0.080	04:10	
Air Blank		04:10	RFI Detect
Air Blank	0.000	04:10	

Data Download Date/Time: 07/26/2017 12:22

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESCOBAR,ALEJANDRO,
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf
ARRESTING OFFICER: ESCOBAR,ALEJANDRO
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf
LAST AGENCY INSPECTION: 06/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 03:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:28

SUBJECT'S LAST NAME: KING

DATE/TIME OF TEST: 07/21/2017 04:15:33

Diagnostics Check		OK	04:12	
Air Blank		0.000	04:13	
Control Test		0.080	04:13	
Air Blank		0.000	04:14	
Subject Sample	# 1		04:14	Subject Test Refused
Breath Volume		0.000	04:14	
Air Blank		0.000	04:14	
Control Test		0.081	04:14	
Air Blank		0.000	04:15	
Diagnostics Check		OK	04:15	

Data Download Date/Time: 07/26/2017 12:22

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORALES,DAVID,J
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf
ARRESTING OFFICER: MORALES,DAVID
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf
LAST AGENCY INSPECTION: 06/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 03:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:25

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 07/21/2017 05:12:56

Diagnostics Check		OK	05:10	
Air Blank		0.000	05:10	
Control Test		0.080	05:10	
Air Blank		0.000	05:11	
Subject Sample	# 1		05:11	Subject Test Refused
Breath Volume		0.000	05:11	
Air Blank		0.000	05:12	
Control Test		0.081	05:12	
Air Blank		0.000	05:12	
Diagnostics Check		OK	05:12	

Data Download Date/Time: 07/26/2017 12:22

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORALES,DAVID,J
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf
ARRESTING OFFICER: MORALES,DAVID
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf
LAST AGENCY INSPECTION: 06/26/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 01:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: PERRAULT

DATE/TIME OF TEST: 07/25/2017 02:56:41

Diagnostics Check		OK	02:53	
Air Blank		0.000	02:54	
Control Test		0.080	02:54	
Air Blank		0.000	02:55	
Subject Sample	# 1		02:55	Subject Test Refused
Breath Volume		0.000	02:55	
Air Blank		0.000	02:55	
Control Test		0.081	02:56	
Air Blank		0.000	02:56	
Diagnostics Check		OK	02:56	

Data Download Date/Time: 07/26/2017 12:22

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESCOBAR,ALEJANDRO,
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf
ARRESTING OFFICER: ESCOBAR,ALEJANDRO
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf
LAST AGENCY INSPECTION: 07/26/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 02:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:57

SUBJECT'S LAST NAME: STEWARD

DATE/TIME OF TEST: 07/30/2017 03:30:34

Diagnostics Check		OK	03:27	
Air Blank		0.000	03:28	
Control Test		0.080	03:28	
Air Blank		0.000	03:29	
Subject Sample	# 1		03:29	Subject Test Refused
Breath Volume		0.000	03:29	
Air Blank		0.000	03:29	
Control Test		0.081	03:29	
Air Blank		0.000	03:30	
Diagnostics Check		OK	03:30	

Data Download Date/Time: 08/30/2017 11:41

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PADRON,RICARDO,
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf
ARRESTING OFFICER: THOMAS,JASON
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf
LAST AGENCY INSPECTION: 07/26/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 06:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:40

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 07/30/2017 07:32:50

Diagnostics Check		OK	07:18	
Air Blank		0.000	07:18	
Control Test		0.081	07:19	
Air Blank		0.000	07:19	
Subject Sample	# 1	0.129	07:22	Volume Not Met
Breath Volume		0.988	07:22	
Air Blank		0.000	07:23	
Air Blank		0.000	07:25	
Subject Sample	# 2	0.127	07:26	
Breath Volume		1.628	07:26	
Air Blank		0.000	07:27	
Air Blank		0.000	07:29	
Subject Sample	# 3	0.134	07:31	
Breath Volume		1.628	07:31	
Air Blank		0.000	07:31	
Control Test		0.079	07:32	
Air Blank		0.000	07:32	
Diagnostics Check		OK	07:32	

Data Download Date/Time: 08/30/2017 11:41

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BYRNE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:22

SUBJECT'S LAST NAME: BRACEWELL

DATE/TIME OF TEST: 07/01/2017 03:46:08

Diagnostics Check		OK	03:42	
Air Blank		0.000	03:43	
Control Test		0.079	03:43	
Air Blank		0.000	03:44	
Subject Sample	# 1		03:44	Subject Test Refused
Breath Volume		0.000	03:44	
Air Blank		0.000	03:45	
Control Test		0.078	03:45	
Air Blank		0.000	03:46	
Diagnostics Check		OK	03:46	

Data Download Date/Time: 07/13/2017 15:07

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPICER,LORETHA,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JONES,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 08:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:30

SUBJECT'S LAST NAME: MCDOWER

DATE/TIME OF TEST: 07/01/2017 10:52:26

Diagnostics Check	OK	10:51	
Air Blank		10:51	Ambient Fail
Air Blank	0.000	10:52	

Data Download Date/Time: 07/13/2017 15:07

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPICER,LORETHA,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JONES,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 08:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:30

SUBJECT'S LAST NAME: MCDOWER

DATE/TIME OF TEST: 07/01/2017 11:34:33

Diagnostics Check		OK	11:31	
Air Blank		0.000	11:32	
Control Test		0.079	11:32	
Air Blank		0.000	11:32	
Subject Sample	# 1		11:33	Subject Test Refused
Breath Volume		0.000	11:33	
Air Blank		0.000	11:33	
Control Test		0.079	11:33	
Air Blank		0.000	11:34	
Diagnostics Check		OK	11:34	

Data Download Date/Time: 07/13/2017 15:07

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIGHTSEY,ROBERT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MILLER,C
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 06/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:48

SUBJECT'S LAST NAME: CAMPBELL

DATE/TIME OF TEST: 07/09/2017 01:15:19

Diagnostics Check		OK	01:08
Air Blank		0.000	01:09
Control Test		0.078	01:09
Air Blank		0.000	01:10
Subject Sample	# 1	0.201	01:10
Breath Volume		2.230	01:10
Air Blank		0.000	01:11
Air Blank		0.000	01:13
Subject Sample	# 2	0.209	01:13
Breath Volume		1.996	01:13
Air Blank		0.000	01:14
Control Test		0.078	01:14
Air Blank		0.000	01:15
Diagnostics Check		OK	01:15

Data Download Date/Time: 07/13/2017 15:07

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIGHTSEY,ROBERT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CAMACHO,K
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 06/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:34

SUBJECT'S LAST NAME: MANALO

DATE/TIME OF TEST: 07/09/2017 03:01:31

Diagnostics Check		OK	02:55
Air Blank		0.000	02:55
Control Test		0.079	02:56
Air Blank		0.000	02:56
Subject Sample	# 1	0.000	02:57
Breath Volume		1.109	02:57
Air Blank		0.000	02:57
Air Blank		0.000	02:59
Subject Sample	# 2	0.000	03:00
Breath Volume		1.261	03:00
Air Blank		0.000	03:00
Control Test		0.079	03:00
Air Blank		0.000	03:01
Diagnostics Check		OK	03:01

Data Download Date/Time: 07/13/2017 15:07

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIGHTSEY,ROBERT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MORRISON,M
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 06/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:41

SUBJECT'S LAST NAME: LEHMAN

DATE/TIME OF TEST: 07/09/2017 05:05:08

Diagnostics Check		OK	05:02	
Air Blank		0.000	05:02	
Control Test		0.078	05:03	
Air Blank		0.000	05:03	
Subject Sample	# 1		05:03	Subject Test Refused
Breath Volume		0.000	05:03	
Air Blank		0.000	05:04	
Control Test		0.078	05:04	
Air Blank		0.000	05:05	
Diagnostics Check		OK	05:05	

Data Download Date/Time: 07/13/2017 15:07

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCOTT,J,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: THOMAS,J
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 07/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: P

DATE/TIME OF TEST: 07/15/2017 03:57:06

Diagnostics Check		OK	03:50
Air Blank		0.000	03:51
Control Test		0.078	03:51
Air Blank		0.000	03:51
Subject Sample	# 1	0.201	03:52
Breath Volume		2.402	03:52
Air Blank		0.000	03:53
Air Blank		0.000	03:54
Subject Sample	# 2	0.209	03:55
Breath Volume		2.847	03:55
Air Blank		0.000	03:56
Control Test		0.078	03:56
Air Blank		0.000	03:57
Diagnostics Check		OK	03:57

Data Download Date/Time: 07/27/2017 11:44

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON,BRADEN,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: STANLEY,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: HENRRIQUEZ

DATE/TIME OF TEST: 07/16/2017 02:31:35

Diagnostics Check		OK	02:25
Air Blank		0.000	02:25
Control Test		0.079	02:25
Air Blank		0.000	02:26
Subject Sample	# 1	0.105	02:27
Breath Volume		1.222	02:27
Air Blank		0.000	02:27
Air Blank		0.000	02:29
Subject Sample	# 2	0.119	02:30
Breath Volume		2.316	02:30
Air Blank		0.000	02:30
Control Test		0.078	02:30
Air Blank		0.000	02:31
Diagnostics Check		OK	02:31

Data Download Date/Time: 07/27/2017 11:44

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBER,JOSEPH,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BROWN,T.A.
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:59

SUBJECT'S LAST NAME: CONE

DATE/TIME OF TEST: 07/22/2017 03:26:26

Diagnostics Check		OK	03:19
Air Blank		0.000	03:20
Control Test		0.078	03:20
Air Blank		0.000	03:21
Subject Sample	# 1	0.030	03:21
Breath Volume		1.332	03:21
Air Blank		0.000	03:22
Air Blank		0.000	03:24
Subject Sample	# 2	0.028	03:24
Breath Volume		1.171	03:24
Air Blank		0.000	03:25
Control Test		0.078	03:25
Air Blank		0.000	03:26
Diagnostics Check		OK	03:26

Data Download Date/Time: 07/27/2017 11:45

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBER,JOSEPH,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FILBERT,E
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:46

SUBJECT'S LAST NAME: OSBORNE

DATE/TIME OF TEST: 07/22/2017 04:14:44

Diagnostics Check		OK	04:08
Air Blank		0.000	04:08
Control Test		0.078	04:09
Air Blank		0.000	04:09
Subject Sample	# 1	0.147	04:10
Breath Volume		2.953	04:10
Air Blank		0.000	04:10
Air Blank		0.000	04:12
Subject Sample	# 2	0.147	04:13
Breath Volume		2.941	04:13
Air Blank		0.000	04:13
Control Test		0.078	04:14
Air Blank		0.000	04:14
Diagnostics Check		OK	04:14

Data Download Date/Time: 07/27/2017 11:45

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBER,JOSEPH,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: STANLEY,S.M.
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:38

SUBJECT'S LAST NAME: VAZQUEZ

DATE/TIME OF TEST: 07/22/2017 05:05:31

Diagnostics Check		OK	04:58
Air Blank		0.000	04:59
Control Test		0.078	04:59
Air Blank		0.000	05:00
Subject Sample	# 1	0.139	05:00
Breath Volume		2.097	05:00
Air Blank		0.000	05:01
Air Blank		0.000	05:03
Subject Sample	# 2	0.131	05:03
Breath Volume		1.546	05:03
Air Blank		0.000	05:04
Control Test		0.076	05:04
Air Blank		0.000	05:05
Diagnostics Check		OK	05:05

Data Download Date/Time: 07/27/2017 11:45

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBER,JOSEPH,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: KERSHAW,S.B.
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:50

SUBJECT'S LAST NAME: HILLIAR

DATE/TIME OF TEST: 07/22/2017 06:16:19

Diagnostics Check		OK	06:10
Air Blank		0.000	06:10
Control Test		0.078	06:10
Air Blank		0.000	06:11
Subject Sample	# 1	0.000	06:11
Breath Volume		1.679	06:11
Air Blank		0.000	06:12
Air Blank		0.000	06:14
Subject Sample	# 2	0.000	06:14
Breath Volume		1.781	06:14
Air Blank		0.000	06:15
Control Test		0.078	06:15
Air Blank		0.000	06:16
Diagnostics Check		OK	06:16

Data Download Date/Time: 07/27/2017 11:45

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FARLEY,J
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 07/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:13

SUBJECT'S LAST NAME: HUTCHESON

DATE/TIME OF TEST: 07/22/2017 23:40:21

Diagnostics Check		OK	23:34
Air Blank		0.000	23:34
Control Test		0.078	23:34
Air Blank		0.000	23:35
Subject Sample	# 1	0.156	23:35
Breath Volume		2.273	23:35
Air Blank		0.000	23:36
Air Blank		0.000	23:38
Subject Sample	# 2	0.160	23:38
Breath Volume		2.167	23:38
Air Blank		0.000	23:39
Control Test		0.078	23:39
Air Blank		0.000	23:40
Diagnostics Check		OK	23:40

Data Download Date/Time: 07/27/2017 11:45

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: GREEN,J
ARRESTING OFFICER AGENCY: Atlantic Beach PD
LAST AGENCY INSPECTION: 07/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:18

SUBJECT'S LAST NAME: JULIANA

DATE/TIME OF TEST: 07/23/2017 01:41:31

Diagnostics Check		OK	01:38	
Air Blank		0.000	01:39	
Control Test		0.079	01:39	
Air Blank		0.000	01:39	
Subject Sample	# 1		01:40	Subject Test Refused
Breath Volume		0.000	01:40	
Air Blank		0.000	01:40	
Control Test		0.078	01:40	
Air Blank		0.000	01:41	
Diagnostics Check		OK	01:41	

Data Download Date/Time: 07/27/2017 11:45

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FARLEY,J
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 07/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:53

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 07/23/2017 02:19:01

Diagnostics Check		OK	02:13
Air Blank		0.000	02:13
Control Test		0.078	02:13
Air Blank		0.000	02:14
Subject Sample	# 1	0.000	02:14
Breath Volume		3.449	02:14
Air Blank		0.000	02:15
Air Blank		0.000	02:17
Subject Sample	# 2	0.000	02:17
Breath Volume		3.136	02:17
Air Blank		0.000	02:18
Control Test		0.079	02:18
Air Blank		0.000	02:18
Diagnostics Check		OK	02:19

Data Download Date/Time: 07/27/2017 11:45

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: GREEN,J
ARRESTING OFFICER AGENCY: Atlantic Beach PD
LAST AGENCY INSPECTION: 07/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:01

SUBJECT'S LAST NAME: INCISO

DATE/TIME OF TEST: 07/23/2017 04:35:26

Diagnostics Check		OK	04:32	
Air Blank		0.000	04:32	
Control Test		0.078	04:32	
Air Blank		0.000	04:33	
Subject Sample	# 1		04:34	Subject Test Refused
Breath Volume		0.000	04:34	
Air Blank		0.000	04:34	
Control Test		0.078	04:34	
Air Blank		0.000	04:35	
Diagnostics Check		OK	04:35	

Data Download Date/Time: 07/27/2017 11:45

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FILBERT,E
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 07/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:03

SUBJECT'S LAST NAME: BARRIOS

DATE/TIME OF TEST: 07/23/2017 05:30:26

Diagnostics Check		OK	05:23
Air Blank		0.000	05:23
Control Test		0.078	05:24
Air Blank		0.000	05:24
Subject Sample	# 1	0.116	05:25
Breath Volume		1.757	05:25
Air Blank		0.000	05:26
Air Blank		0.000	05:28
Subject Sample	# 2	0.124	05:28
Breath Volume		1.117	05:28
Air Blank		0.000	05:29
Control Test		0.078	05:29
Air Blank		0.000	05:30
Diagnostics Check		OK	05:30

Data Download Date/Time: 07/27/2017 11:45

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIGHTSEY,ROBERT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: KERSHAW,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:33

SUBJECT'S LAST NAME: FIERY

DATE/TIME OF TEST: 07/29/2017 02:56:31

Diagnostics Check		OK	02:53	
Air Blank		0.000	02:54	
Control Test		0.079	02:54	
Air Blank		0.000	02:54	
Subject Sample	# 1		02:55	Subject Test Refused
Breath Volume		0.000	02:55	
Air Blank		0.000	02:55	
Control Test		0.078	02:55	
Air Blank		0.000	02:56	
Diagnostics Check		OK	02:56	

Data Download Date/Time: 08/16/2017 12:31

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIGHTSEY,ROBERT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CAMACHO,K
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 07/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:29

SUBJECT'S LAST NAME: JACKS

DATE/TIME OF TEST: 07/29/2017 03:56:01

Diagnostics Check		OK	03:49
Air Blank		0.000	03:50
Control Test		0.079	03:50
Air Blank		0.000	03:51
Subject Sample	# 1	0.113	03:51
Breath Volume		1.257	03:51
Air Blank		0.000	03:52
Air Blank		0.000	03:54
Subject Sample	# 2	0.111	03:54
Breath Volume		1.304	03:54
Air Blank		0.000	03:55
Control Test		0.078	03:55
Air Blank		0.000	03:55
Diagnostics Check		OK	03:56

Data Download Date/Time: 08/16/2017 12:32

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIGHTSEY,ROBERT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SIMMONS,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: SPRATT

DATE/TIME OF TEST: 07/31/2017 03:37:33

Diagnostics Check		OK	03:30
Air Blank		0.000	03:31
Control Test		0.079	03:31
Air Blank		0.000	03:32
Subject Sample	# 1	0.192	03:32
Breath Volume		3.441	03:32
Air Blank		0.000	03:33
Air Blank		0.000	03:35
Subject Sample	# 2	0.177	03:35
Breath Volume		2.394	03:35
Air Blank		0.000	03:36
Control Test		0.078	03:36
Air Blank		0.000	03:37
Diagnostics Check		OK	03:37

Data Download Date/Time: 08/16/2017 12:32

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIGHTSEY,ROBERT,R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: KERSHAW,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:35

SUBJECT'S LAST NAME: DILLAVOU

DATE/TIME OF TEST: 07/31/2017 05:02:32

Diagnostics Check		OK	04:56
Air Blank		0.000	04:56
Control Test		0.079	04:56
Air Blank		0.000	04:57
Subject Sample	# 1	0.086	04:57
Breath Volume		2.046	04:57
Air Blank		0.000	04:58
Air Blank		0.000	05:00
Subject Sample	# 2	0.088	05:00
Breath Volume		2.914	05:00
Air Blank		0.000	05:01
Control Test		0.079	05:01
Air Blank		0.000	05:02
Diagnostics Check		OK	05:02

Data Download Date/Time: 08/16/2017 12:32

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIGHTSEY,ROBERT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BROWN,C
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 07/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:53

SUBJECT'S LAST NAME: MCCRAY

DATE/TIME OF TEST: 07/31/2017 06:20:35

Diagnostics Check		OK	06:13
Air Blank		0.000	06:14
Control Test		0.079	06:14
Air Blank		0.000	06:15
Subject Sample	# 1	0.171	06:15
Breath Volume		1.945	06:15
Air Blank		0.000	06:16
Air Blank		0.000	06:18
Subject Sample	# 2	0.191	06:18
Breath Volume		2.394	06:18
Air Blank		0.000	06:19
Control Test		0.078	06:19
Air Blank		0.000	06:20
Diagnostics Check		OK	06:20

Data Download Date/Time: 08/16/2017 12:32

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARNANDAN,INDROWTIE,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: WILLIAMS,SEAN
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 06/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A3
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: GIPSON

DATE/TIME OF TEST: 07/01/2017 02:24:55

Diagnostics Check		OK	02:18
Air Blank		0.000	02:19
Control Test		0.079	02:19
Air Blank		0.000	02:20
Subject Sample	# 1	0.116	02:20
Breath Volume		2.300	02:20
Air Blank		0.000	02:21
Air Blank		0.000	02:23
Subject Sample	# 2	0.111	02:23
Breath Volume		1.812	02:23
Air Blank		0.000	02:24
Control Test		0.079	02:24
Air Blank		0.000	02:24
Diagnostics Check		OK	02:24

Data Download Date/Time: 07/21/2017 09:46

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SERPHOS,STEVEN
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 06/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A3
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:44

SUBJECT'S LAST NAME: STILES

DATE/TIME OF TEST: 07/02/2017 02:09:11

Diagnostics Check		OK	02:06	
Air Blank		0.000	02:06	
Control Test		0.078	02:07	
Air Blank		0.000	02:07	
Subject Sample	# 1		02:07	Subject Test Refused
Breath Volume		0.000	02:07	
Air Blank		0.000	02:08	
Control Test		0.078	02:08	
Air Blank		0.000	02:09	
Diagnostics Check		OK	02:09	

Data Download Date/Time: 07/21/2017 09:46

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SERPHOS,STEVEN
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 06/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A3
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:44

SUBJECT'S LAST NAME: STILES

DATE/TIME OF TEST: 07/02/2017 02:17:46

Diagnostics Check		OK	02:10
Air Blank		0.000	02:11
Control Test		0.078	02:11
Air Blank		0.000	02:12
Subject Sample	# 1	0.208	02:13
Breath Volume		2.128	02:13
Air Blank		0.000	02:13
Air Blank		0.000	02:15
Subject Sample	# 2	0.218	02:16
Breath Volume		2.082	02:16
Air Blank		0.000	02:16
Control Test		0.079	02:17
Air Blank		0.000	02:17
Diagnostics Check		OK	02:17

Data Download Date/Time: 07/21/2017 09:46

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: VAN TIL,ETHAN
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 06/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A3
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 14:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:48

SUBJECT'S LAST NAME: SIMON INTERIANO

DATE/TIME OF TEST: 07/02/2017 16:21:52

Diagnostics Check		OK	16:15
Air Blank		0.000	16:15
Control Test		0.078	16:16
Air Blank		0.000	16:16
Subject Sample	# 1	0.218	16:17
Breath Volume		1.847	16:17
Air Blank		0.000	16:17
Air Blank		0.000	16:19
Subject Sample	# 2	0.226	16:20
Breath Volume		3.027	16:20
Air Blank		0.000	16:20
Control Test		0.079	16:21
Air Blank		0.000	16:21
Diagnostics Check		OK	16:21

Data Download Date/Time: 07/21/2017 09:46

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FLINK,ANDREW
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 06/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A3
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 18:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:53

SUBJECT'S LAST NAME: LAWLOR

DATE/TIME OF TEST: 07/02/2017 20:19:33

Diagnostics Check		OK	20:13
Air Blank		0.000	20:13
Control Test		0.079	20:14
Air Blank		0.000	20:14
Subject Sample	# 1	0.000	20:15
Breath Volume		1.914	20:15
Air Blank		0.000	20:15
Air Blank		0.000	20:17
Subject Sample	# 2	0.000	20:18
Breath Volume		1.195	20:18
Air Blank		0.000	20:18
Control Test		0.079	20:18
Air Blank		0.000	20:19
Diagnostics Check		OK	20:19

Data Download Date/Time: 07/21/2017 09:46

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINA,MARMILLY,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: THOMAZ,ANTONIO
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 06/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A3
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 07/04/2017 02:22:23

Diagnostics Check		OK	02:15
Air Blank		0.000	02:16
Control Test		0.078	02:16
Air Blank		0.000	02:16
Subject Sample	# 1	0.226	02:17
Breath Volume		1.902	02:17
Air Blank		0.000	02:18
Air Blank		0.000	02:20
Subject Sample	# 2	0.232	02:20
Breath Volume		2.457	02:20
Air Blank		0.000	02:21
Control Test		0.078	02:21
Air Blank		0.000	02:22
Diagnostics Check		OK	02:22

Data Download Date/Time: 07/21/2017 09:46

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: OLEA,ELIEZER
ARRESTING OFFICER AGENCY: Pembroke Pines PD
LAST AGENCY INSPECTION: 06/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A3
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 08:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:45

SUBJECT'S LAST NAME: STEPHENS

DATE/TIME OF TEST: 07/04/2017 10:12:07

Diagnostics Check		OK	10:05
Air Blank		0.000	10:06
Control Test		0.079	10:06
Air Blank		0.000	10:07
Subject Sample	# 1	0.164	10:07
Breath Volume		1.867	10:07
Air Blank		0.000	10:08
Air Blank		0.000	10:10
Subject Sample	# 2	0.163	10:10
Breath Volume		1.503	10:10
Air Blank		0.000	10:11
Control Test		0.079	10:11
Air Blank		0.000	10:12
Diagnostics Check		OK	10:12

Data Download Date/Time: 07/21/2017 09:46

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SAPP,JOSHUA
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 06/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A3
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: GIRALDO

DATE/TIME OF TEST: 07/05/2017 02:44:47

Diagnostics Check		OK	02:38
Air Blank		0.000	02:38
Control Test		0.079	02:39
Air Blank		0.000	02:39
Subject Sample	# 1	0.193	02:40
Breath Volume		2.386	02:40
Air Blank		0.000	02:40
Air Blank		0.000	02:42
Subject Sample	# 2	0.186	02:43
Breath Volume		1.871	02:43
Air Blank		0.000	02:43
Control Test		0.078	02:44
Air Blank		0.000	02:44
Diagnostics Check		OK	02:44

Data Download Date/Time: 07/21/2017 09:46

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON,YVETTE,D
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FOWLER,SHAUN
ARRESTING OFFICER AGENCY: Hollywood PD
LAST AGENCY INSPECTION: 06/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A3
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 12:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:10

SUBJECT'S LAST NAME: MIRANDA

DATE/TIME OF TEST: 07/05/2017 13:56:33

Diagnostics Check		OK	13:45	
Air Blank		0.000	13:45	
Control Test		0.079	13:45	
Air Blank		0.000	13:46	
Subject Sample	# 1		13:49	No Sample Provided
Breath Volume		0.000	13:49	
Air Blank		0.000	13:50	
Air Blank		0.000	13:52	
Subject Sample	# 2	0.000	13:55	Volume Not Met
Breath Volume		0.304	13:55	
Air Blank		0.000	13:55	
Control Test		0.080	13:55	
Air Blank		0.000	13:56	
Diagnostics Check		OK	13:56	

Data Download Date/Time: 07/21/2017 09:46

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON,YVETTE,D
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FOWLER,SHAUN
ARRESTING OFFICER AGENCY: Hollywood PD
LAST AGENCY INSPECTION: 06/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A3
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 12:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:10

SUBJECT'S LAST NAME: MIRANDA

DATE/TIME OF TEST: 07/05/2017 14:06:32

Diagnostics Check		OK	13:59
Air Blank		0.000	13:59
Control Test		0.079	14:00
Air Blank		0.000	14:00
Subject Sample	# 1	0.000	14:02
Breath Volume		1.640	14:02
Air Blank		0.000	14:02
Air Blank		0.000	14:04
Subject Sample	# 2	0.000	14:05
Breath Volume		1.832	14:05
Air Blank		0.000	14:05
Control Test		0.079	14:05
Air Blank		0.000	14:06
Diagnostics Check		OK	14:06

Data Download Date/Time: 07/21/2017 09:46

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SAPP,JOSHUA
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 06/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A3
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:04

SUBJECT'S LAST NAME: AYERS

DATE/TIME OF TEST: 07/05/2017 22:31:47

Diagnostics Check		OK	22:25
Air Blank		0.000	22:25
Control Test		0.079	22:25
Air Blank		0.000	22:26
Subject Sample	# 1	0.225	22:27
Breath Volume		1.667	22:27
Air Blank		0.000	22:28
Air Blank		0.000	22:29
Subject Sample	# 2	0.212	22:30
Breath Volume		1.109	22:30
Air Blank		0.000	22:30
Control Test		0.078	22:31
Air Blank		0.000	22:31
Diagnostics Check		OK	22:31

Data Download Date/Time: 07/21/2017 09:46

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: WILIAMS,SEAN
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 06/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A3
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:52

SUBJECT'S LAST NAME: BARTOLON BARTOLON

DATE/TIME OF TEST: 07/08/2017 04:22:42

Diagnostics Check		OK	04:15
Air Blank		0.000	04:16
Control Test		0.079	04:16
Air Blank		0.000	04:17
Subject Sample	# 1	0.367	04:17
Breath Volume		2.042	04:17
Air Blank		0.000	04:18
Air Blank		0.000	04:19
Subject Sample	# 2	0.355	04:21
Breath Volume		2.328	04:21
Air Blank		0.000	04:21
Control Test		0.078	04:22
Air Blank		0.000	04:22
Diagnostics Check		OK	04:22

Data Download Date/Time: 07/21/2017 09:46

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON,YVETTE,D
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SAPP,JOSHUA
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 06/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A3
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 16:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:05

SUBJECT'S LAST NAME: PICKETT

DATE/TIME OF TEST: 07/13/2017 17:40:56

Diagnostics Check		OK	17:33
Air Blank		0.000	17:34
Control Test		0.079	17:34
Air Blank		0.000	17:34
Subject Sample	# 1	0.311	17:36
Breath Volume		1.160	17:36
Air Blank		0.000	17:36
Air Blank		0.000	17:38
Subject Sample	# 2	0.309	17:39
Breath Volume		2.179	17:39
Air Blank		0.000	17:40
Control Test		0.078	17:40
Air Blank		0.000	17:40
Diagnostics Check		OK	17:40

Data Download Date/Time: 07/21/2017 09:46

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: PLANTE,MILTON
ARRESTING OFFICER AGENCY: Sunrise PD
LAST AGENCY INSPECTION: 06/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A3
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:11

SUBJECT'S LAST NAME: LOMBARDI

DATE/TIME OF TEST: 07/14/2017 23:41:31

Diagnostics Check		OK	23:34
Air Blank		0.000	23:34
Control Test		0.079	23:34
Air Blank		0.000	23:35
Subject Sample	# 1	0.322	23:36
Breath Volume		1.566	23:36
Air Blank		0.000	23:36
Air Blank		0.000	23:38
Subject Sample	# 2	0.320	23:40
Breath Volume		1.804	23:40
Air Blank		0.000	23:40
Control Test		0.078	23:40
Air Blank		0.000	23:41
Diagnostics Check		OK	23:41

Data Download Date/Time: 07/21/2017 09:46

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINA,MARMILLY,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: MENDEZ,NELSON
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 06/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A3
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 07/16/2017 03:40:58

Diagnostics Check		OK	03:34
Air Blank		0.000	03:34
Control Test		0.079	03:34
Air Blank		0.000	03:35
Subject Sample	# 1	0.197	03:36
Breath Volume		1.906	03:36
Air Blank		0.000	03:36
Air Blank		0.000	03:38
Subject Sample	# 2	0.185	03:39
Breath Volume		1.527	03:39
Air Blank		0.000	03:40
Control Test		0.079	03:40
Air Blank		0.000	03:40
Diagnostics Check		OK	03:40

Data Download Date/Time: 07/21/2017 09:46

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARNANDAN,INDROWTIE,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GARCIA,HEBERTO
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 06/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A3
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 08:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:20

SUBJECT'S LAST NAME: MATA GARCIA

DATE/TIME OF TEST: 07/16/2017 10:03:48

Diagnostics Check		OK	09:50	
Air Blank		0.000	09:50	
Control Test		0.078	09:51	
Air Blank		0.000	09:51	
Subject Sample	# 1		09:54	No Sample Provided
Breath Volume		0.000	09:54	
Air Blank		0.000	09:55	
Air Blank		0.000	09:57	
Subject Sample	# 2	0.246	09:57	
Breath Volume		2.386	09:57	
Air Blank		0.000	09:58	
Air Blank		0.000	10:00	
Subject Sample	# 3	0.258	10:02	
Breath Volume		1.320	10:02	
Air Blank		0.000	10:02	
Control Test		0.079	10:03	
Air Blank		0.000	10:03	
Diagnostics Check		OK	10:03	

Data Download Date/Time: 07/21/2017 09:46

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: JACKSON,K
ARRESTING OFFICER AGENCY: Plantation PD
LAST AGENCY INSPECTION: 06/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A3
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 17:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:10

SUBJECT'S LAST NAME: LAROCHE

DATE/TIME OF TEST: 07/16/2017 21:36:32

Diagnostics Check		OK	21:30
Air Blank		0.000	21:30
Control Test		0.078	21:31
Air Blank		0.000	21:31
Subject Sample	# 1	0.110	21:32
Breath Volume		3.265	21:32
Air Blank		0.000	21:32
Air Blank		0.000	21:34
Subject Sample	# 2	0.110	21:35
Breath Volume		3.062	21:35
Air Blank		0.000	21:35
Control Test		0.078	21:35
Air Blank		0.000	21:36
Diagnostics Check		OK	21:36

Data Download Date/Time: 07/21/2017 09:46

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINA,MARMILLY,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: LUNA,KRISTINA
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 06/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A3
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: JEANTY

DATE/TIME OF TEST: 07/16/2017 23:34:16

Diagnostics Check		OK	23:30	
Air Blank		0.000	23:30	
Control Test		0.079	23:31	
Air Blank		0.000	23:31	
Subject Sample	# 1		23:32	Subject Test Refused
Breath Volume		0.000	23:32	
Air Blank		0.000	23:33	
Control Test		0.078	23:33	
Air Blank		0.000	23:34	
Diagnostics Check		OK	23:34	

Data Download Date/Time: 07/21/2017 09:46

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GARCIA,MELANIE
ARRESTING OFFICER AGENCY: Hollywood PD
LAST AGENCY INSPECTION: 07/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A3
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:11

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 07/21/2017 21:53:47

Diagnostics Check		OK	21:47
Air Blank		0.000	21:47
Control Test		0.079	21:48
Air Blank		0.000	21:48
Subject Sample	# 1	0.179	21:49
Breath Volume		2.003	21:49
Air Blank		0.000	21:49
Air Blank		0.000	21:51
Subject Sample	# 2	0.182	21:52
Breath Volume		1.363	21:52
Air Blank		0.000	21:52
Control Test		0.080	21:53
Air Blank		0.000	21:53
Diagnostics Check		OK	21:53

Data Download Date/Time: 08/28/2017 15:57

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARNANDAN,INDROWTIE,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: DALTON,RYAN
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 07/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A3
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: STOCKARD

DATE/TIME OF TEST: 07/22/2017 03:17:39

Diagnostics Check		OK	03:10
Air Blank		0.000	03:10
Control Test		0.079	03:11
Air Blank		0.000	03:11
Subject Sample	# 1	0.123	03:13
Breath Volume		1.339	03:13
Air Blank		0.000	03:13
Air Blank		0.000	03:15
Subject Sample	# 2	0.140	03:16
Breath Volume		1.679	03:16
Air Blank		0.000	03:16
Control Test		0.078	03:17
Air Blank		0.000	03:17
Diagnostics Check		OK	03:17

Data Download Date/Time: 08/28/2017 15:57

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FITHIAN,JAY
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 07/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A3
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:02

SUBJECT'S LAST NAME: QUINTERO-BRANCACHO

DATE/TIME OF TEST: 07/22/2017 04:34:05

Diagnostics Check		OK	04:28
Air Blank		0.000	04:28
Control Test		0.078	04:28
Air Blank		0.000	04:29
Subject Sample	# 1	0.179	04:29
Breath Volume		1.207	04:29
Air Blank		0.000	04:30
Air Blank		0.000	04:32
Subject Sample	# 2	0.172	04:32
Breath Volume		1.863	04:32
Air Blank		0.000	04:33
Control Test		0.079	04:33
Air Blank		0.000	04:33
Diagnostics Check		OK	04:34

Data Download Date/Time: 08/28/2017 15:57

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINA,MARMILLY,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: NAVARRO,CAZIMIRRO
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 07/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A3
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: MORGAN

DATE/TIME OF TEST: 07/23/2017 04:16:15

Diagnostics Check		OK	04:08
Air Blank		0.000	04:09
Control Test		0.078	04:09
Air Blank		0.000	04:09
Subject Sample	# 1	0.132	04:11
Breath Volume		2.417	04:11
Air Blank		0.000	04:12
Air Blank		0.000	04:14
Subject Sample	# 2	0.137	04:14
Breath Volume		2.347	04:14
Air Blank		0.000	04:15
Control Test		0.078	04:15
Air Blank		0.000	04:16
Diagnostics Check		OK	04:16

Data Download Date/Time: 08/28/2017 15:57

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINA,MARMILLY,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: MCMAHON,AUSTIN
ARRESTING OFFICER AGENCY: Sunrise PD
LAST AGENCY INSPECTION: 07/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A3
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 06:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:08

SUBJECT'S LAST NAME: ABAD VASQUEZ

DATE/TIME OF TEST: 07/23/2017 07:52:27

Diagnostics Check		OK	07:45
Air Blank		0.000	07:46
Control Test		0.079	07:46
Air Blank		0.000	07:46
Subject Sample	# 1	0.141	07:48
Breath Volume		2.097	07:48
Air Blank		0.000	07:48
Air Blank		0.000	07:50
Subject Sample	# 2	0.140	07:51
Breath Volume		1.605	07:51
Air Blank		0.000	07:51
Control Test		0.078	07:51
Air Blank		0.000	07:52
Diagnostics Check		OK	07:52

Data Download Date/Time: 08/28/2017 15:57

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINA,MARMILLY,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: PACK,JEFFREY
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 07/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A3
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:37

SUBJECT'S LAST NAME: FAULKNOR

DATE/TIME OF TEST: 07/24/2017 05:19:32

Diagnostics Check		OK	05:12
Air Blank		0.000	05:12
Control Test		0.078	05:12
Air Blank		0.000	05:13
Subject Sample	# 1	0.202	05:14
Breath Volume		1.773	05:14
Air Blank		0.000	05:15
Air Blank		0.000	05:17
Subject Sample	# 2	0.205	05:18
Breath Volume		1.875	05:18
Air Blank		0.000	05:18
Control Test		0.078	05:18
Air Blank		0.000	05:19
Diagnostics Check		OK	05:19

Data Download Date/Time: 08/28/2017 15:57

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: DEFEO,PATRICK
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 07/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A3
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: STANN

DATE/TIME OF TEST: 07/28/2017 22:58:01

Diagnostics Check		OK	22:51
Air Blank		0.000	22:51
Control Test		0.079	22:52
Air Blank		0.000	22:52
Subject Sample	# 1	0.155	22:53
Breath Volume		1.191	22:53
Air Blank		0.000	22:53
Air Blank		0.000	22:55
Subject Sample	# 2	0.155	22:56
Breath Volume		1.269	22:56
Air Blank		0.000	22:57
Control Test		0.079	22:57
Air Blank		0.000	22:57
Diagnostics Check		OK	22:58

Data Download Date/Time: 08/28/2017 15:57

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARNANDAN,INDROWTIE,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: KELLY,TAYLOR
ARRESTING OFFICER AGENCY: Hallandale Beach PD
LAST AGENCY INSPECTION: 07/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A3
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:35

SUBJECT'S LAST NAME: GUERRA

DATE/TIME OF TEST: 07/29/2017 05:02:27

Diagnostics Check		OK	04:55
Air Blank		0.000	04:55
Control Test		0.080	04:56
Air Blank		0.000	04:56
Subject Sample	# 1	0.171	04:57
Breath Volume		1.792	04:57
Air Blank		0.000	04:58
Air Blank		0.000	05:00
Subject Sample	# 2	0.169	05:01
Breath Volume		1.609	05:01
Air Blank		0.000	05:01
Control Test		0.079	05:01
Air Blank		0.000	05:02
Diagnostics Check		OK	05:02

Data Download Date/Time: 08/28/2017 15:57

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARNANDAN,INDROWTIE,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: DE LOS REYES,YOAN
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 07/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A3
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 07:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:45

SUBJECT'S LAST NAME: BADERE

DATE/TIME OF TEST: 07/29/2017 08:11:53

Diagnostics Check		OK	08:05
Air Blank		0.000	08:05
Control Test		0.079	08:06
Air Blank		0.000	08:06
Subject Sample	# 1	0.036	08:07
Breath Volume		1.554	08:07
Air Blank		0.000	08:07
Air Blank		0.000	08:09
Subject Sample	# 2	0.036	08:10
Breath Volume		1.707	08:10
Air Blank		0.000	08:10
Control Test		0.079	08:11
Air Blank		0.000	08:11
Diagnostics Check		OK	08:11

Data Download Date/Time: 08/28/2017 15:57

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINA,MARMILLY,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: TRABUE,THOMAS
ARRESTING OFFICER AGENCY: Pembroke Pines PD
LAST AGENCY INSPECTION: 07/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A3
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 05:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:30

SUBJECT'S LAST NAME: DIAZ ROJAS

DATE/TIME OF TEST: 07/30/2017 07:29:20

Diagnostics Check		OK	07:21
Air Blank		0.000	07:22
Control Test		0.078	07:22
Air Blank		0.000	07:22
Subject Sample	# 1	0.211	07:24
Breath Volume		1.777	07:24
Air Blank		0.000	07:24
Air Blank		0.000	07:26
Subject Sample	# 2	0.193	07:27
Breath Volume		1.453	07:27
Air Blank		0.000	07:28
Control Test		0.079	07:28
Air Blank		0.000	07:29
Diagnostics Check		OK	07:29

Data Download Date/Time: 08/28/2017 15:57

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINA,MARMILLY,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: O'DOR,ROBERT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 07/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A3
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: GAYDA

DATE/TIME OF TEST: 07/30/2017 22:12:36

Diagnostics Check		OK	22:06
Air Blank		0.000	22:06
Control Test		0.079	22:07
Air Blank		0.000	22:07
Subject Sample	# 1	0.180	22:08
Breath Volume		1.160	22:08
Air Blank		0.000	22:08
Air Blank		0.000	22:10
Subject Sample	# 2	0.178	22:11
Breath Volume		1.945	22:11
Air Blank		0.000	22:11
Control Test		0.079	22:12
Air Blank		0.000	22:12
Diagnostics Check		OK	22:12

Data Download Date/Time: 08/28/2017 15:57

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001050

INSTRUMENT REGISTERED TO: CORAL SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WETHERINGTON,STEPHEN,L
BREATH TEST OPERATOR AGENCY: Coral Springs PD
ARRESTING OFFICER: WETHERINGTON,STEPHEN
ARRESTING OFFICER AGENCY: Coral Springs PD
LAST AGENCY INSPECTION: 06/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A2
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:59

SUBJECT'S LAST NAME: GARAICOA GYARMATI

DATE/TIME OF TEST: 07/09/2017 05:54:08

Diagnostics Check		OK	05:47
Air Blank		0.000	05:47
Control Test		0.080	05:48
Air Blank		0.000	05:48
Subject Sample	# 1	0.138	05:49
Breath Volume		3.644	05:49
Air Blank		0.000	05:49
Air Blank		0.000	05:51
Subject Sample	# 2	0.138	05:52
Breath Volume		2.359	05:52
Air Blank		0.000	05:53
Control Test		0.080	05:53
Air Blank		0.000	05:54
Diagnostics Check		OK	05:54

Data Download Date/Time: 07/19/2017 15:28

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001050

INSTRUMENT REGISTERED TO: CORAL SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MURDOCK,TROY,J
BREATH TEST OPERATOR AGENCY: Coral Springs PD
ARRESTING OFFICER: MURDOCK,TROY
ARRESTING OFFICER AGENCY: Coral Springs PD
LAST AGENCY INSPECTION: 06/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A2
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 05:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:42

SUBJECT'S LAST NAME: BRANNON

DATE/TIME OF TEST: 07/16/2017 06:21:00

Diagnostics Check		OK	06:14
Air Blank		0.000	06:15
Control Test		0.080	06:15
Air Blank		0.000	06:16
Subject Sample	# 1	0.000	06:16
Breath Volume		2.726	06:16
Air Blank		0.000	06:17
Air Blank		0.000	06:19
Subject Sample	# 2	0.000	06:19
Breath Volume		3.199	06:19
Air Blank		0.000	06:20
Control Test		0.080	06:20
Air Blank		0.000	06:20
Diagnostics Check		OK	06:21

Data Download Date/Time: 07/19/2017 15:28

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001050

INSTRUMENT REGISTERED TO: CORAL SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WETHERINGTON,STEPHEN,L
BREATH TEST OPERATOR AGENCY: Coral Springs PD
ARRESTING OFFICER: WETHERINGTON,STEPHEN
ARRESTING OFFICER AGENCY: Coral Springs PD
LAST AGENCY INSPECTION: 06/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A2
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:51

SUBJECT'S LAST NAME: BARRETT-THOMPSON

DATE/TIME OF TEST: 07/18/2017 21:26:06

Diagnostics Check		OK	21:18
Air Blank		0.000	21:18
Control Test		0.080	21:18
Air Blank		0.000	21:19
Subject Sample	# 1	0.259	21:20
Breath Volume		2.773	21:20
Air Blank		0.000	21:21
Air Blank		0.000	21:23
Subject Sample	# 2	0.257	21:24
Breath Volume		2.484	21:24
Air Blank		0.000	21:25
Control Test		0.079	21:25
Air Blank		0.000	21:26
Diagnostics Check		OK	21:26

Data Download Date/Time: 07/19/2017 15:28

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001050

INSTRUMENT REGISTERED TO: CORAL SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WETHERINGTON,STEPHEN,L
BREATH TEST OPERATOR AGENCY: Coral Springs PD
ARRESTING OFFICER: WETHERINGTON,STEPHEN
ARRESTING OFFICER AGENCY: Coral Springs PD
LAST AGENCY INSPECTION: 07/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:51

SUBJECT'S LAST NAME: PIERRE

DATE/TIME OF TEST: 07/22/2017 02:25:31

Diagnostics Check		OK	02:18
Air Blank		0.000	02:19
Control Test		0.081	02:19
Air Blank		0.000	02:19
Subject Sample	# 1	0.214	02:20
Breath Volume		2.792	02:20
Air Blank		0.000	02:21
Air Blank		0.000	02:23
Subject Sample	# 2	0.228	02:23
Breath Volume		1.867	02:23
Air Blank		0.000	02:24
Control Test		0.080	02:24
Air Blank		0.000	02:25
Diagnostics Check		OK	02:25

Data Download Date/Time: 08/29/2017 14:02

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001050

INSTRUMENT REGISTERED TO: CORAL SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WETHERINGTON,STEPHEN,L
BREATH TEST OPERATOR AGENCY: Coral Springs PD
ARRESTING OFFICER: WETHERINGTON,STEPHEN
ARRESTING OFFICER AGENCY: Coral Springs PD
LAST AGENCY INSPECTION: 07/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:42

SUBJECT'S LAST NAME: TOLEDO MAZUERA

DATE/TIME OF TEST: 07/27/2017 05:18:30

Diagnostics Check		OK	05:12
Air Blank		0.000	05:12
Control Test		0.081	05:12
Air Blank		0.000	05:13
Subject Sample	# 1	0.110	05:13
Breath Volume		1.949	05:13
Air Blank		0.000	05:14
Air Blank		0.000	05:16
Subject Sample	# 2	0.110	05:16
Breath Volume		2.179	05:16
Air Blank		0.000	05:17
Control Test		0.081	05:17
Air Blank		0.000	05:18
Diagnostics Check		OK	05:18

Data Download Date/Time: 08/29/2017 14:02

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001050

INSTRUMENT REGISTERED TO: CORAL SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRAWFORD,DANIELLE,A
BREATH TEST OPERATOR AGENCY: Coral Springs PD
ARRESTING OFFICER: KENNY,MICHAEL
ARRESTING OFFICER AGENCY: Coral Springs PD
LAST AGENCY INSPECTION: 07/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 05:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:45

SUBJECT'S LAST NAME: TOLA

DATE/TIME OF TEST: 07/30/2017 06:26:06

Diagnostics Check		OK	06:18
Air Blank		0.000	06:19
Control Test		0.081	06:19
Air Blank		0.000	06:20
Subject Sample	# 1	0.243	06:21
Breath Volume		2.316	06:21
Air Blank		0.000	06:21
Air Blank		0.000	06:23
Subject Sample	# 2	0.232	06:24
Breath Volume		2.531	06:24
Air Blank		0.000	06:25
Control Test		0.080	06:25
Air Blank		0.000	06:26
Diagnostics Check		OK	06:26

Data Download Date/Time: 08/29/2017 14:02

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001050

INSTRUMENT REGISTERED TO: CORAL SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BYRNE,GRETCHEN,
BREATH TEST OPERATOR AGENCY: Coral Springs PD
ARRESTING OFFICER: BYRNE,GRETCHEN
ARRESTING OFFICER AGENCY: Coral Springs PD
LAST AGENCY INSPECTION: 07/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 09:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:10

SUBJECT'S LAST NAME: CAPABLANCA

DATE/TIME OF TEST: 07/30/2017 10:54:57

Diagnostics Check		OK	10:43	
Air Blank		0.000	10:44	
Control Test		0.080	10:44	
Air Blank		0.000	10:44	
Subject Sample	# 1		10:47	No Sample Provided
Breath Volume		0.000	10:47	
Air Blank		0.000	10:48	
Air Blank		0.000	10:50	
Subject Sample	# 2		10:53	No Sample Provided
Breath Volume		0.000	10:53	
Air Blank		0.000	10:54	
Control Test		0.081	10:54	
Air Blank		0.000	10:54	
Diagnostics Check		OK	10:54	

Data Download Date/Time: 08/29/2017 14:02

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001050

INSTRUMENT REGISTERED TO: CORAL SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BYRNE,GRETCHEN,
BREATH TEST OPERATOR AGENCY: Coral Springs PD
ARRESTING OFFICER: BYRNE,GRETCHEN
ARRESTING OFFICER AGENCY: Coral Springs PD
LAST AGENCY INSPECTION: 07/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 09:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:50

SUBJECT'S LAST NAME: CAPABLANCA

DATE/TIME OF TEST: 07/30/2017 11:15:25

Diagnostics Check		OK	11:10	
Air Blank		0.000	11:10	
Control Test		0.080	11:11	
Air Blank		0.000	11:11	
Subject Sample	# 1		11:14	Subject Test Refused
Breath Volume		0.000	11:14	
Air Blank		0.000	11:14	
Control Test		0.081	11:14	
Air Blank		0.000	11:15	
Diagnostics Check		OK	11:15	

Data Download Date/Time: 08/29/2017 14:02

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001050

INSTRUMENT REGISTERED TO: CORAL SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BYRNE,GRETCHEN,
BREATH TEST OPERATOR AGENCY: Coral Springs PD
ARRESTING OFFICER: BYRNE,GRETCHEN
ARRESTING OFFICER AGENCY: Coral Springs PD
LAST AGENCY INSPECTION: 07/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 09:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:50

SUBJECT'S LAST NAME: CAPABLANCA

DATE/TIME OF TEST: 07/30/2017 11:21:07

Diagnostics Check		OK	11:18	
Air Blank		0.000	11:18	
Control Test		0.081	11:19	
Air Blank		0.000	11:19	
Subject Sample	# 1		11:19	Subject Test Refused
Breath Volume		0.000	11:19	
Air Blank		0.000	11:20	
Control Test		0.080	11:20	
Air Blank		0.000	11:21	
Diagnostics Check		OK	11:21	

Data Download Date/Time: 08/29/2017 14:02

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALDO,AARON,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ALLI,ANTHONY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: MARIK

DATE/TIME OF TEST: 07/01/2017 01:27:53

Diagnostics Check		OK	01:21
Air Blank		0.000	01:22
Control Test		0.080	01:22
Air Blank		0.000	01:22
Subject Sample	# 1	0.149	01:23
Breath Volume		3.136	01:23
Air Blank		0.000	01:24
Air Blank		0.000	01:25
Subject Sample	# 2	0.146	01:26
Breath Volume		3.218	01:26
Air Blank		0.000	01:26
Control Test		0.079	01:27
Air Blank		0.000	01:27
Diagnostics Check		OK	01:27

Data Download Date/Time: 07/04/2017 16:12

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLI,ANTHONY,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: WALDO,AARON
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:49

SUBJECT'S LAST NAME: STONE

DATE/TIME OF TEST: 07/02/2017 02:16:36

Diagnostics Check		OK	02:09
Air Blank		0.000	02:10
Control Test		0.080	02:10
Air Blank		0.000	02:11
Subject Sample	# 1	0.172	02:11
Breath Volume		4.511	02:11
Air Blank		0.000	02:12
Air Blank		0.000	02:14
Subject Sample	# 2	0.181	02:15
Breath Volume		4.476	02:15
Air Blank		0.000	02:15
Control Test		0.079	02:16
Air Blank		0.000	02:16
Diagnostics Check		OK	02:16

Data Download Date/Time: 07/04/2017 16:12

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEISKOPF,MICHAEL,D
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: LAMOUR,ZACHARY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 07/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 17:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:37

SUBJECT'S LAST NAME: SZANTO

DATE/TIME OF TEST: 07/06/2017 18:09:41

Diagnostics Check		OK	18:03
Air Blank		0.000	18:04
Control Test		0.079	18:04
Air Blank		0.000	18:04
Subject Sample	# 1	0.211	18:05
Breath Volume		2.863	18:05
Air Blank		0.000	18:05
Air Blank		0.000	18:07
Subject Sample	# 2	0.213	18:08
Breath Volume		2.367	18:08
Air Blank		0.000	18:08
Control Test		0.079	18:09
Air Blank		0.000	18:09
Diagnostics Check		OK	18:09

Data Download Date/Time: 08/05/2017 17:28

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALDO,AARON,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ALLI,ANTHONY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 07/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:55

SUBJECT'S LAST NAME: CALABRESE

DATE/TIME OF TEST: 07/06/2017 20:24:47

Diagnostics Check		OK	20:15
Air Blank		0.000	20:16
Control Test		0.079	20:16
Air Blank		0.000	20:17
Subject Sample	# 1	0.255	20:20
Breath Volume		1.578	20:20
Air Blank		0.000	20:20
Air Blank		0.000	20:22
Subject Sample	# 2	0.237	20:23
Breath Volume		2.078	20:23
Air Blank		0.000	20:23
Control Test		0.078	20:24
Air Blank		0.000	20:24
Diagnostics Check		OK	20:24

Data Download Date/Time: 08/05/2017 17:28

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALDO,AARON,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: WEISKOPF,MICHAEL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 07/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:33

SUBJECT'S LAST NAME: DEROSIERS

DATE/TIME OF TEST: 07/07/2017 23:59:10

Diagnostics Check		OK	23:53	
Air Blank		0.000	23:54	
Control Test		0.080	23:54	
Air Blank		0.000	23:54	
Subject Sample	# 1		23:57	Subject Test Refused
Breath Volume		0.000	23:57	
Air Blank		0.000	23:58	
Control Test		0.080	23:58	
Air Blank		0.000	23:59	
Diagnostics Check		OK	23:59	

Data Download Date/Time: 08/05/2017 17:28

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KARAYIANES,MICHAEL,F
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ZOLLOTA,VLADIMIR
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 07/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:26

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 07/08/2017 02:53:23

Diagnostics Check		OK	02:47
Air Blank		0.000	02:47
Control Test		0.080	02:48
Air Blank		0.000	02:48
Subject Sample	# 1	0.000	02:49
Breath Volume		2.292	02:49
Air Blank		0.000	02:49
Air Blank		0.000	02:51
Subject Sample	# 2	0.000	02:52
Breath Volume		2.523	02:52
Air Blank		0.000	02:52
Control Test		0.079	02:52
Air Blank		0.000	02:53
Diagnostics Check		OK	02:53

Data Download Date/Time: 08/05/2017 17:28

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZOLLOTA,VLADIMIR,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: COX,LENARD
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 07/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:35

SUBJECT'S LAST NAME: ROSSEY

DATE/TIME OF TEST: 07/08/2017 21:02:16

Diagnostics Check		OK	20:55
Air Blank		0.000	20:56
Control Test		0.079	20:56
Air Blank		0.000	20:56
Subject Sample	# 1	0.000	20:57
Breath Volume		3.039	20:57
Air Blank		0.000	20:58
Air Blank		0.000	21:00
Subject Sample	# 2	0.000	21:00
Breath Volume		2.937	21:00
Air Blank		0.000	21:01
Control Test		0.080	21:01
Air Blank		0.000	21:02
Diagnostics Check		OK	21:02

Data Download Date/Time: 08/05/2017 17:28

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,EMILY,K
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: SMITH,EMILY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 07/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 17:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:23

SUBJECT'S LAST NAME: LASSO

DATE/TIME OF TEST: 07/09/2017 17:48:53

Diagnostics Check		OK	17:43	
Air Blank		0.000	17:43	
Control Test		0.079	17:44	
Air Blank		0.000	17:44	
Subject Sample	# 1	0.171	17:45	
Breath Volume		4.167	17:45	
Air Blank		0.000	17:45	
Air Blank		0.000	17:47	
Subject Sample	# 2	0.169	17:48	
Breath Volume		3.656	17:48	
Air Blank			17:48	RFI Detect
Air Blank		0.000	17:48	

Data Download Date/Time: 08/05/2017 17:28

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALDO,AARON,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: WEISKOPF,MICHAEL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 07/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:01

SUBJECT'S LAST NAME: HOWES

DATE/TIME OF TEST: 07/14/2017 00:23:33

Diagnostics Check		OK	00:21	
Air Blank		0.000	00:21	
Control Test		0.080	00:22	
Air Blank		0.000	00:22	
Subject Sample	# 1		00:23	Sequence Aborted
Breath Volume		0.000	00:23	
Air Blank		0.000	00:23	

Data Download Date/Time: 08/05/2017 17:28

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALDO,AARON,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: WEISKOPF,MICHAEL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 07/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: HOWES

DATE/TIME OF TEST: 07/14/2017 00:50:43

Diagnostics Check		OK	00:47	
Air Blank		0.000	00:48	
Control Test		0.080	00:48	
Air Blank		0.000	00:49	
Subject Sample	# 1		00:49	Subject Test Refused
Breath Volume		0.000	00:49	
Air Blank		0.000	00:49	
Control Test		0.080	00:50	
Air Blank		0.000	00:50	
Diagnostics Check		OK	00:50	

Data Download Date/Time: 08/05/2017 17:28

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ACRI,MICHEAL,N
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: SAIA,MICHAEL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 07/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 05:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:50

SUBJECT'S LAST NAME: PORRAS

DATE/TIME OF TEST: 07/16/2017 06:16:57

Diagnostics Check		OK	06:10
Air Blank		0.000	06:11
Control Test		0.079	06:11
Air Blank		0.000	06:11
Subject Sample	# 1	0.179	06:12
Breath Volume		3.203	06:12
Air Blank		0.000	06:13
Air Blank		0.000	06:15
Subject Sample	# 2	0.175	06:15
Breath Volume		3.656	06:15
Air Blank		0.000	06:16
Control Test		0.080	06:16
Air Blank		0.000	06:16
Diagnostics Check		OK	06:16

Data Download Date/Time: 08/05/2017 17:28

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZOLLOTA,VLADIMIR,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: COX,LENARD
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 07/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:05

SUBJECT'S LAST NAME: UBER

DATE/TIME OF TEST: 07/20/2017 23:29:08

Diagnostics Check		OK	23:25	
Air Blank		0.000	23:26	
Control Test		0.079	23:26	
Air Blank		0.000	23:26	
Subject Sample	# 1		23:27	RFI Detect
Breath Volume		0.355	23:27	
Air Blank		0.000	23:28	
Control Test		0.080	23:28	
Air Blank		0.000	23:29	
Diagnostics Check		OK	23:29	

Data Download Date/Time: 08/05/2017 17:28

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZOLLOTA,VLADIMIR,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: COX,LENARD
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 07/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:05

SUBJECT'S LAST NAME: UBER

DATE/TIME OF TEST: 07/20/2017 23:38:27

Diagnostics Check		OK	23:30	
Air Blank		0.000	23:30	
Control Test		0.079	23:31	
Air Blank		0.000	23:31	
Subject Sample	# 1	0.197	23:34	
Breath Volume		1.390	23:34	
Air Blank		0.000	23:34	
Air Blank		0.000	23:36	
Subject Sample	# 2	0.185	23:37	
Breath Volume		1.234	23:37	
Air Blank		0.000	23:37	
Control Test		0.077	23:37	RFI Detect
Air Blank		0.000	23:38	

Data Download Date/Time: 08/05/2017 17:28

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZOLLOTA,VLADIMIR,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: COX,LENARD
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 07/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:05

SUBJECT'S LAST NAME: UBER

DATE/TIME OF TEST: 07/20/2017 23:49:46

Diagnostics Check		OK	23:42
Air Blank		0.000	23:42
Control Test		0.079	23:43
Air Blank		0.000	23:43
Subject Sample	# 1	0.202	23:44
Breath Volume		1.312	23:44
Air Blank		0.000	23:45
Air Blank		0.000	23:47
Subject Sample	# 2	0.196	23:48
Breath Volume		1.382	23:48
Air Blank		0.000	23:48
Control Test		0.077	23:49
Air Blank		0.000	23:49
Diagnostics Check		OK	23:49

Data Download Date/Time: 08/05/2017 17:28

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALDO,AARON,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: WEISKOPF,MICHAEL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 07/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:53

SUBJECT'S LAST NAME: CURP

DATE/TIME OF TEST: 07/21/2017 02:23:09

Diagnostics Check		OK	02:13
Air Blank		0.000	02:14
Control Test		0.079	02:14
Air Blank		0.000	02:15
Subject Sample	# 1	0.156	02:17
Breath Volume		1.496	02:17
Air Blank		0.000	02:18
Air Blank		0.000	02:20
Subject Sample	# 2	0.158	02:21
Breath Volume		2.140	02:21
Air Blank		0.000	02:22
Control Test		0.079	02:22
Air Blank		0.000	02:23
Diagnostics Check		OK	02:23

Data Download Date/Time: 08/05/2017 17:28

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEISKOPF,MICHAEL,D
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: WALDO,AARON
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 07/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: TILLER

DATE/TIME OF TEST: 07/23/2017 03:39:51

Diagnostics Check		OK	03:36	
Air Blank		0.000	03:36	
Control Test		0.079	03:37	
Air Blank		0.000	03:37	
Subject Sample	# 1		03:38	Subject Test Refused
Breath Volume		0.000	03:38	
Air Blank		0.000	03:38	
Control Test		0.078	03:39	
Air Blank		0.000	03:39	
Diagnostics Check		OK	03:39	

Data Download Date/Time: 08/05/2017 17:28

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,EMILY,K
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: SMITH,EMILY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 07/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 16:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:30

SUBJECT'S LAST NAME: LAWRENCE

DATE/TIME OF TEST: 07/23/2017 17:53:21

Diagnostics Check		OK	17:50	
Air Blank		0.000	17:50	
Control Test		0.079	17:51	
Air Blank		0.000	17:51	
Subject Sample	# 1		17:51	Subject Test Refused
Breath Volume		0.000	17:51	
Air Blank		0.000	17:52	
Control Test		0.079	17:52	
Air Blank		0.000	17:53	
Diagnostics Check		OK	17:53	

Data Download Date/Time: 08/05/2017 17:29

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEISKOPF,MICHAEL,D
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ALLI,ANTHONY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 07/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:48

SUBJECT'S LAST NAME: PLEASANTS

DATE/TIME OF TEST: 07/27/2017 02:15:13

Diagnostics Check		OK	02:08
Air Blank		0.000	02:09
Control Test		0.080	02:09
Air Blank		0.000	02:09
Subject Sample	# 1	0.150	02:10
Breath Volume		3.472	02:10
Air Blank		0.000	02:11
Air Blank		0.000	02:13
Subject Sample	# 2	0.150	02:13
Breath Volume		2.996	02:13
Air Blank		0.000	02:14
Control Test		0.079	02:14
Air Blank		0.000	02:15
Diagnostics Check		OK	02:15

Data Download Date/Time: 08/05/2017 17:29

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLI,ANTHONY,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: KARAYIANES,MICHAEL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 07/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: TUCKER

DATE/TIME OF TEST: 07/27/2017 02:59:34

Diagnostics Check		OK	02:52
Air Blank		0.000	02:53
Control Test		0.080	02:53
Air Blank		0.000	02:54
Subject Sample	# 1	0.000	02:55
Breath Volume		4.179	02:55
Air Blank		0.000	02:55
Air Blank		0.000	02:57
Subject Sample	# 2	0.000	02:58
Breath Volume		3.269	02:58
Air Blank		0.000	02:58
Control Test		0.080	02:58
Air Blank		0.000	02:59
Diagnostics Check		OK	02:59

Data Download Date/Time: 08/05/2017 17:29

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLI,ANTHONY,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: WEISKOPF,MICHAEL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 07/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:36

SUBJECT'S LAST NAME: DONOVAN

DATE/TIME OF TEST: 07/27/2017 04:07:58

Diagnostics Check		OK	03:56	
Air Blank		0.000	03:56	
Control Test		0.080	03:57	
Air Blank		0.000	03:57	
Subject Sample	# 1		04:00	No Sample Provided
Breath Volume		0.000	04:00	
Air Blank		0.000	04:01	
Air Blank		0.000	04:03	
Subject Sample	# 2	0.189	04:06	Volume Not Met
Breath Volume		0.445	04:06	
Air Blank		0.000	04:07	
Control Test		0.080	04:07	
Air Blank		0.000	04:07	
Diagnostics Check		OK	04:07	

Data Download Date/Time: 08/05/2017 17:29

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLI,ANTHONY,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: WEISKOPF,MICHAEL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 07/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:36

SUBJECT'S LAST NAME: DONOVAN

DATE/TIME OF TEST: 07/27/2017 04:21:55

Diagnostics Check		OK	04:09	
Air Blank		0.000	04:09	
Control Test		0.080	04:10	
Air Blank		0.000	04:10	
Subject Sample	# 1	0.205	04:13	Volume Not Met
Breath Volume		0.414	04:13	
Air Blank		0.000	04:14	
Air Blank		0.000	04:16	
Subject Sample	# 2	0.219	04:16	No .020 agreement
Breath Volume		1.507	04:16	
Air Blank		0.000	04:17	
Air Blank		0.000	04:19	
Subject Sample	# 3	0.177	04:20	No .020 agreement
Breath Volume		1.347	04:20	
Air Blank		0.000	04:20	
Control Test		0.079	04:21	
Air Blank		0.000	04:21	
Diagnostics Check		OK	04:21	

Data Download Date/Time: 08/05/2017 17:29

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLI,ANTHONY,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: WEISKOPF,MICHAEL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 07/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:36

SUBJECT'S LAST NAME: DONOVAN

DATE/TIME OF TEST: 07/27/2017 04:38:11

Diagnostics Check		OK	04:23	
Air Blank		0.000	04:24	
Control Test		0.079	04:24	
Air Blank		0.000	04:25	
Subject Sample	# 1	0.220	04:27	
Breath Volume		1.355	04:27	
Air Blank		0.000	04:27	
Air Blank		0.000	04:29	
Subject Sample	# 2	0.175	04:32	Volume Not Met
Breath Volume		0.484	04:32	
Air Blank		0.000	04:33	
Air Blank		0.000	04:35	
Subject Sample	# 3	0.211	04:36	
Breath Volume		1.429	04:36	
Air Blank		0.000	04:37	
Control Test		0.078	04:37	
Air Blank		0.000	04:38	
Diagnostics Check		OK	04:38	

Data Download Date/Time: 08/05/2017 17:29

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPBELL,D,R
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: GORDON,G
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 06/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A3
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:23

SUBJECT'S LAST NAME: LIVINGSTONE

DATE/TIME OF TEST: 07/01/2017 04:46:58

Diagnostics Check		OK	04:44	
Air Blank		0.000	04:44	
Control Test		0.078	04:44	
Air Blank		0.000	04:45	
Subject Sample	# 1		04:45	Subject Test Refused
Breath Volume		0.000	04:45	
Air Blank		0.000	04:46	
Control Test		0.080	04:46	
Air Blank		0.000	04:46	
Diagnostics Check		OK	04:46	

Data Download Date/Time: 07/25/2017 07:52

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THIRSK,KEVIN,M
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: GOETHE,R
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 06/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A3
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 19:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:20

SUBJECT'S LAST NAME: FLORES ALMANZA

DATE/TIME OF TEST: 07/01/2017 20:41:30

Diagnostics Check	OK	20:40	
Air Blank	0.000	20:40	
Control Test	0.079	20:40	Interferent Detect
Air Blank	0.000	20:41	

Data Download Date/Time: 07/25/2017 07:52

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THIRSK,KEVIN,M
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: GOETHE,R
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 06/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A3
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 19:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:20

SUBJECT'S LAST NAME: FLORES ALMANZA

DATE/TIME OF TEST: 07/01/2017 20:44:49

Diagnostics Check	OK	20:43	
Air Blank	0.000	20:43	
Control Test	0.079	20:44	Interferent Detect
Air Blank	0.000	20:44	

Data Download Date/Time: 07/25/2017 07:52

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THIRSK,KEVIN,M
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: GOETHE,R
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 06/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A3
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 19:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:20

SUBJECT'S LAST NAME: FLORES ALMANZA

DATE/TIME OF TEST: 07/01/2017 21:03:44

Diagnostics Check	OK	21:02	
Air Blank	0.000	21:02	
Control Test	0.079	21:03	Interferent Detect
Air Blank	0.000	21:03	

Data Download Date/Time: 07/25/2017 07:52

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THIRSK,KEVIN,M
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: GOETHE,R
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 06/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A3
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 19:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:20

SUBJECT'S LAST NAME: FLORES ALMANZA

DATE/TIME OF TEST: 07/01/2017 21:09:08

Diagnostics Check	OK	21:07	
Air Blank	0.000	21:08	
Control Test	0.079	21:08	Interferent Detect
Air Blank	0.000	21:09	

Data Download Date/Time: 07/25/2017 07:52

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THIRSK,KEVIN,M
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: GOETHE,R
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 06/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A3
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 19:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:20

SUBJECT'S LAST NAME: FLORES ALMANZA

DATE/TIME OF TEST: 07/01/2017 21:19:20

Diagnostics Check	OK	21:18	
Air Blank	0.000	21:18	
Control Test	0.079	21:18	Interferent Detect
Air Blank	0.000	21:19	

Data Download Date/Time: 07/25/2017 07:52

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCNEIL,K,P
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: TEST,TEST
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 06/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A3
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: ADMIN
ARREST TIME: 23:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: TEST

DATE/TIME OF TEST: 07/02/2017 00:34:20

Diagnostics Check	OK	00:33	
Air Blank	0.000	00:33	
Control Test	0.080	00:33	Interferent Detect
Air Blank	0.000	00:34	

Data Download Date/Time: 07/25/2017 07:52

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THIRSK,KEVIN,M
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: THIRSK,K
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 06/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A3
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: ADMIN
ARREST TIME: 00:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: TEST

DATE/TIME OF TEST: 07/02/2017 01:17:54

Diagnostics Check	OK	01:16	
Air Blank	0.000	01:17	
Control Test	0.080	01:17	Interferent Detect
Air Blank	0.000	01:17	

Data Download Date/Time: 07/25/2017 07:52

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THIRSK,KEVIN,M
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: THIRSK,K
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 06/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A3
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: ADMIN
ARREST TIME: 01:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: TEST

DATE/TIME OF TEST: 07/02/2017 02:06:47

Diagnostics Check	OK	02:05	
Air Blank	0.000	02:05	
Control Test	0.076	02:06	Interferent Detect
Air Blank	0.000	02:06	

Data Download Date/Time: 07/25/2017 07:52

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THIRSK,KEVIN,M
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: THIRSK,K
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 06/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A3
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: ADMIN
ARREST TIME: 18:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:40

SUBJECT'S LAST NAME: TEST

DATE/TIME OF TEST: 07/02/2017 19:06:40

Diagnostics Check		OK	19:00
Air Blank		0.000	19:01
Control Test		0.079	19:01
Air Blank		0.000	19:01
Subject Sample	# 1	0.000	19:02
Breath Volume		2.160	19:02
Air Blank		0.000	19:02
Air Blank		0.000	19:04
Subject Sample	# 2	0.000	19:05
Breath Volume		2.519	19:05
Air Blank		0.000	19:05
Control Test		0.079	19:06
Air Blank		0.000	19:06
Diagnostics Check		OK	19:06

Data Download Date/Time: 07/25/2017 07:52

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HEALY,GREGORY,B
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: HEALY,GREGORY
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 06/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A3
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: PERRY

DATE/TIME OF TEST: 07/04/2017 22:11:07

Diagnostics Check		OK	22:05
Air Blank		0.000	22:05
Control Test		0.079	22:05
Air Blank		0.000	22:06
Subject Sample	# 1	0.163	22:06
Breath Volume		3.011	22:06
Air Blank		0.000	22:07
Air Blank		0.000	22:09
Subject Sample	# 2	0.162	22:09
Breath Volume		3.750	22:09
Air Blank		0.000	22:10
Control Test		0.078	22:10
Air Blank		0.000	22:11
Diagnostics Check		OK	22:11

Data Download Date/Time: 07/25/2017 07:52

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPBELL,D,R
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: REYNOLDS,M
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 06/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A3
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: CHAPMAN

DATE/TIME OF TEST: 07/07/2017 04:08:33

Diagnostics Check		OK	04:01
Air Blank		0.000	04:02
Control Test		0.079	04:02
Air Blank		0.000	04:03
Subject Sample	# 1	0.250	04:03
Breath Volume		1.113	04:03
Air Blank		0.000	04:04
Air Blank		0.000	04:06
Subject Sample	# 2	0.235	04:07
Breath Volume		1.453	04:07
Air Blank		0.000	04:07
Control Test		0.077	04:07
Air Blank		0.000	04:08
Diagnostics Check		OK	04:08

Data Download Date/Time: 07/25/2017 07:52

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GILL,T,V
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: KNIGHT,G
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 06/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A3
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:53

SUBJECT'S LAST NAME: NELSON

DATE/TIME OF TEST: 07/07/2017 22:20:54

Diagnostics Check		OK	22:14
Air Blank		0.000	22:14
Control Test		0.080	22:15
Air Blank		0.000	22:15
Subject Sample	# 1	0.157	22:16
Breath Volume		2.906	22:16
Air Blank		0.000	22:16
Air Blank		0.000	22:18
Subject Sample	# 2	0.166	22:19
Breath Volume		3.332	22:19
Air Blank		0.000	22:20
Control Test		0.079	22:20
Air Blank		0.000	22:20
Diagnostics Check		OK	22:20

Data Download Date/Time: 07/25/2017 07:52

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCNEIL,K,P
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: REYNOLDS,M
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 06/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A3
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:51

SUBJECT'S LAST NAME: STEVENS

DATE/TIME OF TEST: 07/11/2017 00:18:31

Diagnostics Check		OK	00:11
Air Blank		0.000	00:11
Control Test		0.080	00:12
Air Blank		0.000	00:12
Subject Sample	# 1	0.194	00:13
Breath Volume		1.703	00:13
Air Blank		0.000	00:14
Air Blank		0.000	00:16
Subject Sample	# 2	0.196	00:17
Breath Volume		1.832	00:17
Air Blank		0.000	00:17
Control Test		0.079	00:17
Air Blank		0.000	00:18
Diagnostics Check		OK	00:18

Data Download Date/Time: 07/25/2017 07:52

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EDGY,JOY,F
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: GRAY,D
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 06/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A3
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 14:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:07

SUBJECT'S LAST NAME: PEDRERA

DATE/TIME OF TEST: 07/12/2017 16:36:55

Diagnostics Check		OK	16:31
Air Blank		0.000	16:31
Control Test		0.080	16:31
Air Blank		0.000	16:32
Subject Sample	# 1	0.000	16:32
Breath Volume		4.144	16:32
Air Blank		0.000	16:33
Air Blank		0.000	16:35
Subject Sample	# 2	0.000	16:35
Breath Volume		3.425	16:35
Air Blank		0.000	16:36
Control Test		0.080	16:36
Air Blank		0.000	16:36
Diagnostics Check		OK	16:36

Data Download Date/Time: 07/25/2017 07:52

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HEALY,G,B
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: HEALY,G
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 06/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A3
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: GRANGER

DATE/TIME OF TEST: 07/16/2017 03:38:30

Diagnostics Check		OK	03:35	
Air Blank		0.000	03:36	
Control Test		0.079	03:36	
Air Blank		0.000	03:36	
Subject Sample	# 1		03:37	Subject Test Refused
Breath Volume		0.000	03:37	
Air Blank		0.000	03:37	
Control Test		0.079	03:37	
Air Blank		0.000	03:38	
Diagnostics Check		OK	03:38	

Data Download Date/Time: 07/25/2017 07:52

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCNEIL,K,P
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: FENDER,A
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 06/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A3
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:11

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 07/16/2017 04:34:01

Diagnostics Check		OK	04:31	
Air Blank		0.000	04:31	
Control Test		0.080	04:32	
Air Blank		0.000	04:32	
Subject Sample	# 1		04:32	Subject Test Refused
Breath Volume		0.000	04:32	
Air Blank		0.000	04:33	
Control Test		0.079	04:33	
Air Blank		0.000	04:33	
Diagnostics Check		OK	04:34	

Data Download Date/Time: 07/25/2017 07:52

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GILL,T,V
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: CONE,A
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 06/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A3
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: WELSH

DATE/TIME OF TEST: 07/24/2017 04:11:02

Diagnostics Check		OK	04:08	
Air Blank		0.000	04:08	
Control Test		0.080	04:09	
Air Blank		0.000	04:09	
Subject Sample	# 1		04:09	Subject Test Refused
Breath Volume		0.000	04:09	
Air Blank		0.000	04:10	
Control Test		0.080	04:10	
Air Blank		0.000	04:10	
Diagnostics Check		OK	04:11	

Data Download Date/Time: 07/25/2017 07:52

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMMONDS,TEAQULIA,L
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: JOGLAR,JOSE
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 06/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: COX

DATE/TIME OF TEST: 07/01/2017 02:35:01

Diagnostics Check		OK	02:27
Air Blank		0.000	02:27
Control Test		0.078	02:28
Air Blank		0.000	02:28
Subject Sample	# 1	0.000	02:30
Breath Volume		1.429	02:30
Air Blank		0.000	02:31
Air Blank		0.000	02:33
Subject Sample	# 2	0.000	02:33
Breath Volume		1.542	02:33
Air Blank		0.000	02:34
Control Test		0.077	02:34
Air Blank		0.000	02:34
Diagnostics Check		OK	02:35

Data Download Date/Time: 07/07/2017 09:38

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POTTER,ADAM,H
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: POTTER,ADAM
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 06/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:28

SUBJECT'S LAST NAME: GODINEZ

DATE/TIME OF TEST: 07/01/2017 21:14:58

Diagnostics Check		OK	21:12	
Air Blank		0.000	21:12	
Control Test		0.078	21:12	
Air Blank		0.000	21:13	
Subject Sample	# 1		21:13	Subject Test Refused
Breath Volume		0.000	21:13	
Air Blank		0.000	21:14	
Control Test		0.078	21:14	
Air Blank		0.000	21:14	
Diagnostics Check		OK	21:14	

Data Download Date/Time: 07/07/2017 09:38

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COSTALES,ALEXANDRO,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: COSTALES,ALEXANDRO
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:57

SUBJECT'S LAST NAME: ROSS

DATE/TIME OF TEST: 07/01/2017 22:34:43

Diagnostics Check		OK	22:28
Air Blank		0.000	22:29
Control Test		0.078	22:29
Air Blank		0.000	22:29
Subject Sample	# 1	0.000	22:30
Breath Volume		2.226	22:30
Air Blank		0.000	22:30
Air Blank		0.000	22:32
Subject Sample	# 2	0.000	22:33
Breath Volume		1.863	22:33
Air Blank		0.000	22:33
Control Test		0.078	22:34
Air Blank		0.000	22:34
Diagnostics Check		OK	22:34

Data Download Date/Time: 07/07/2017 09:38

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROOKS,JOSEPH,P
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)
ARRESTING OFFICER: BROOKS,JOSEPH
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 06/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: BUI
ARREST TIME: 15:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:13

SUBJECT'S LAST NAME: SITEK

DATE/TIME OF TEST: 07/02/2017 16:41:41

Diagnostics Check		OK	16:38	
Air Blank		0.000	16:39	
Control Test		0.078	16:39	
Air Blank		0.000	16:40	
Subject Sample	# 1		16:40	Subject Test Refused
Breath Volume		0.000	16:40	
Air Blank		0.000	16:40	
Control Test		0.079	16:41	
Air Blank		0.000	16:41	
Diagnostics Check		OK	16:41	

Data Download Date/Time: 07/07/2017 09:38

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARREN,BILLY,J
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: WARREN,BILLY
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 06/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 19:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:56

SUBJECT'S LAST NAME: DELGADO

DATE/TIME OF TEST: 07/02/2017 20:29:16

Diagnostics Check		OK	20:22
Air Blank		0.000	20:23
Control Test		0.078	20:23
Air Blank		0.000	20:24
Subject Sample	# 1	0.173	20:24
Breath Volume		2.214	20:24
Air Blank		0.000	20:25
Air Blank		0.000	20:27
Subject Sample	# 2	0.163	20:27
Breath Volume		1.632	20:27
Air Blank		0.000	20:28
Control Test		0.078	20:28
Air Blank		0.000	20:29
Diagnostics Check		OK	20:29

Data Download Date/Time: 07/07/2017 09:38

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCGOUGH,ADRIN,D
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: EARLS,CHRISTIAN
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 06/06/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: WINDHAM

DATE/TIME OF TEST: 07/03/2017 00:43:10

Diagnostics Check		OK	00:40	
Air Blank		0.000	00:40	
Control Test		0.078	00:40	
Air Blank		0.000	00:41	
Subject Sample	# 1		00:41	Subject Test Refused
Breath Volume		0.000	00:41	
Air Blank		0.000	00:42	
Control Test		0.077	00:42	
Air Blank		0.000	00:43	
Diagnostics Check		OK	00:43	

Data Download Date/Time: 07/07/2017 09:38

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCGOUGH,ADRIN,D
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: WALDORFF,RUSSELL
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 06:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:30

SUBJECT'S LAST NAME: BAILEY

DATE/TIME OF TEST: 07/03/2017 07:03:35

Diagnostics Check		OK	06:57
Air Blank		0.000	06:57
Control Test		0.078	06:58
Air Blank		0.000	06:58
Subject Sample	# 1	0.134	06:59
Breath Volume		1.816	06:59
Air Blank		0.000	06:59
Air Blank		0.000	07:01
Subject Sample	# 2	0.139	07:02
Breath Volume		1.718	07:02
Air Blank		0.000	07:02
Control Test		0.079	07:02
Air Blank		0.000	07:03
Diagnostics Check		OK	07:03

Data Download Date/Time: 07/07/2017 09:38

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUCAS,NATHANIEL,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: CARTER,JACOB
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 06/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: MAYA

DATE/TIME OF TEST: 07/04/2017 02:03:44

Diagnostics Check		OK	02:00	
Air Blank		0.000	02:01	
Control Test		0.078	02:01	
Air Blank		0.000	02:02	
Subject Sample	# 1		02:02	Subject Test Refused
Breath Volume		0.000	02:02	
Air Blank		0.000	02:02	
Control Test		0.078	02:03	
Air Blank		0.000	02:03	
Diagnostics Check		OK	02:03	

Data Download Date/Time: 07/07/2017 09:38

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAWES,ERIC,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: HERRERA,ARTURO
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 07/07/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: QUEIROZ

DATE/TIME OF TEST: 07/08/2017 00:59:28

Diagnostics Check		OK	00:52
Air Blank		0.000	00:53
Control Test		0.078	00:53
Air Blank		0.000	00:54
Subject Sample	# 1	0.188	00:55
Breath Volume		1.515	00:55
Air Blank		0.000	00:55
Air Blank		0.000	00:57
Subject Sample	# 2	0.192	00:57
Breath Volume		1.761	00:57
Air Blank		0.000	00:58
Control Test		0.078	00:58
Air Blank		0.000	00:59
Diagnostics Check		OK	00:59

Data Download Date/Time: 08/07/2017 14:59

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAWES,ERIC,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: WOFFORD,CLINT
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 07/07/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:52

SUBJECT'S LAST NAME: NOEL

DATE/TIME OF TEST: 07/08/2017 01:27:35

Diagnostics Check		OK	01:21
Air Blank		0.000	01:21
Control Test		0.078	01:21
Air Blank		0.000	01:22
Subject Sample	# 1	0.096	01:23
Breath Volume		2.640	01:23
Air Blank		0.000	01:23
Air Blank		0.000	01:25
Subject Sample	# 2	0.095	01:26
Breath Volume		2.890	01:26
Air Blank		0.000	01:26
Control Test		0.077	01:26
Air Blank		0.000	01:27
Diagnostics Check		OK	01:27

Data Download Date/Time: 08/07/2017 14:59

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUCAS,NATHANIEL,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: LUCAS,NATHANIEL
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 07/07/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: FULLER

DATE/TIME OF TEST: 07/09/2017 03:34:04

Diagnostics Check		OK	03:27
Air Blank		0.000	03:27
Control Test		0.079	03:28
Air Blank		0.000	03:28
Subject Sample	# 1	0.154	03:29
Breath Volume		1.425	03:29
Air Blank		0.000	03:29
Air Blank		0.000	03:31
Subject Sample	# 2	0.156	03:32
Breath Volume		1.933	03:32
Air Blank		0.000	03:33
Control Test		0.079	03:33
Air Blank		0.000	03:33
Diagnostics Check		OK	03:34

Data Download Date/Time: 08/07/2017 14:59

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HASENZAHL,STEPHEN,B
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: HASENZAHL,STEPHEN
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 07/07/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: CASTILLO

DATE/TIME OF TEST: 07/09/2017 04:39:49

Diagnostics Check		OK	04:26	
Air Blank		0.000	04:27	
Control Test		0.078	04:27	
Air Blank		0.000	04:28	
Subject Sample	# 1	0.169	04:31	Volume Not Met
Breath Volume		0.503	04:31	
Air Blank		0.000	04:31	
Air Blank		0.000	04:33	
Subject Sample	# 2	0.212	04:35	
Breath Volume		1.886	04:35	
Air Blank		0.000	04:35	
Air Blank		0.000	04:37	
Subject Sample	# 3	0.205	04:38	
Breath Volume		1.445	04:38	
Air Blank		0.000	04:38	
Control Test		0.077	04:39	
Air Blank		0.000	04:39	
Diagnostics Check		OK	04:39	

Data Download Date/Time: 08/07/2017 14:59

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOOHER,JOY,S
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: WILE,JESSICA
ARRESTING OFFICER AGENCY: Winter Haven PD
LAST AGENCY INSPECTION: 07/07/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: CLARK

DATE/TIME OF TEST: 07/10/2017 01:51:06

Diagnostics Check		OK	01:48	
Air Blank		0.000	01:48	
Control Test		0.078	01:49	
Air Blank		0.000	01:49	
Subject Sample	# 1		01:49	Subject Test Refused
Breath Volume		0.000	01:49	
Air Blank		0.000	01:50	
Control Test		0.079	01:50	
Air Blank		0.000	01:51	
Diagnostics Check		OK	01:51	

Data Download Date/Time: 08/07/2017 14:59

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OGG,DAVID,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: DANIEL,JAMES
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 07/07/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 14:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:24

SUBJECT'S LAST NAME: REYES

DATE/TIME OF TEST: 07/11/2017 14:57:15

Diagnostics Check		OK	14:49
Air Blank		0.000	14:50
Control Test		0.078	14:50
Air Blank		0.000	14:50
Subject Sample	# 1	0.313	14:52
Breath Volume		1.394	14:52
Air Blank		0.000	14:52
Air Blank		0.000	14:54
Subject Sample	# 2	0.300	14:55
Breath Volume		1.546	14:55
Air Blank		0.000	14:56
Control Test		0.078	14:56
Air Blank		0.000	14:57
Diagnostics Check		OK	14:57

Data Download Date/Time: 08/07/2017 14:59

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAWES,ERIC,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: LAWES,ERIC
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 07/07/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: PADILLA-CUBILLO

DATE/TIME OF TEST: 07/12/2017 04:49:06

Diagnostics Check		OK	04:37
Air Blank		0.000	04:37
Control Test		0.078	04:38
Air Blank		0.000	04:38
Subject Sample	# 1	0.192	04:40
Breath Volume		2.316	04:40
Air Blank		0.000	04:41
Air Blank		0.000	04:43
Subject Sample	# 2	0.216	04:44
Breath Volume		2.574	04:44
Air Blank		0.000	04:44
Air Blank		0.000	04:46
Subject Sample	# 3	0.215	04:47
Breath Volume		1.699	04:47
Air Blank		0.000	04:48
Control Test		0.077	04:48
Air Blank		0.000	04:49
Diagnostics Check		OK	04:49

Data Download Date/Time: 08/07/2017 14:59

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORA,AARON,J
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: MORA,AARON
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/07/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 10:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:31

SUBJECT'S LAST NAME: GABLER

DATE/TIME OF TEST: 07/13/2017 11:58:40

Diagnostics Check		OK	11:52
Air Blank		0.000	11:52
Control Test		0.078	11:53
Air Blank		0.000	11:53
Subject Sample	# 1	0.000	11:54
Breath Volume		3.246	11:54
Air Blank		0.000	11:54
Air Blank		0.000	11:56
Subject Sample	# 2	0.000	11:57
Breath Volume		3.035	11:57
Air Blank		0.000	11:57
Control Test		0.077	11:58
Air Blank		0.000	11:58
Diagnostics Check		OK	11:58

Data Download Date/Time: 08/07/2017 14:59

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SYKES,STEVEN,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: SYKES,STEVEN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 07/07/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:30

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 07/13/2017 21:10:31

Diagnostics Check		OK	20:56	
Air Blank		0.000	20:57	
Control Test		0.078	20:57	
Air Blank		0.000	20:57	
Subject Sample	# 1	0.120	21:00	
Breath Volume		1.546	21:00	
Air Blank		0.000	21:00	
Air Blank		0.000	21:02	
Subject Sample	# 2	0.113	21:05	Volume Not Met
Breath Volume		0.687	21:05	
Air Blank		0.000	21:06	
Air Blank		0.000	21:08	
Subject Sample	# 3	0.126	21:09	
Breath Volume		2.261	21:09	
Air Blank		0.000	21:09	
Control Test		0.077	21:09	
Air Blank		0.000	21:10	
Diagnostics Check		OK	21:10	

Data Download Date/Time: 08/07/2017 14:59

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUCAS,NATHANIEL,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: LUCAS,NATHANIEL
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 07/07/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:05

SUBJECT'S LAST NAME: SOLDATKIN

DATE/TIME OF TEST: 07/13/2017 21:36:42

Diagnostics Check		OK	21:30
Air Blank		0.000	21:31
Control Test		0.078	21:31
Air Blank		0.000	21:31
Subject Sample	# 1	0.364	21:32
Breath Volume		2.339	21:32
Air Blank		0.000	21:32
Air Blank		0.000	21:34
Subject Sample	# 2	0.368	21:35
Breath Volume		2.449	21:35
Air Blank		0.000	21:35
Control Test		0.078	21:36
Air Blank		0.000	21:36
Diagnostics Check		OK	21:36

Data Download Date/Time: 08/07/2017 14:59

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOOHER,JOY,S
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: WOFFORD,CLINT
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 07/07/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:27

SUBJECT'S LAST NAME: CROCK

DATE/TIME OF TEST: 07/14/2017 01:09:35

Diagnostics Check		OK	01:02
Air Blank		0.000	01:03
Control Test		0.078	01:03
Air Blank		0.000	01:03
Subject Sample	# 1	0.245	01:04
Breath Volume		2.843	01:04
Air Blank		0.000	01:05
Air Blank		0.000	01:07
Subject Sample	# 2	0.254	01:08
Breath Volume		3.066	01:08
Air Blank		0.000	01:08
Control Test		0.078	01:09
Air Blank		0.000	01:09
Diagnostics Check		OK	01:09

Data Download Date/Time: 08/07/2017 14:59

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PITTS,JAMES,J
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: DUNN,CHRISTOPHER
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 07/07/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 14:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:55

SUBJECT'S LAST NAME: MERCER

DATE/TIME OF TEST: 07/15/2017 15:27:41

Diagnostics Check		OK	15:21
Air Blank		0.000	15:21
Control Test		0.079	15:22
Air Blank		0.000	15:22
Subject Sample	# 1	0.000	15:23
Breath Volume		2.722	15:23
Air Blank		0.000	15:23
Air Blank		0.000	15:25
Subject Sample	# 2	0.000	15:26
Breath Volume		1.867	15:26
Air Blank		0.000	15:26
Control Test		0.079	15:27
Air Blank		0.000	15:27
Diagnostics Check		OK	15:27

Data Download Date/Time: 08/07/2017 14:59

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTLE,CORY,A
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: SUTTLE,CORY
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 07/07/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: VELCO

DATE/TIME OF TEST: 07/18/2017 03:39:44

Diagnostics Check		OK	03:30
Air Blank		0.000	03:31
Control Test		0.078	03:31
Air Blank		0.000	03:32
Subject Sample	# 1	0.151	03:32
Breath Volume		2.121	03:32
Air Blank		0.000	03:33
Air Blank		0.000	03:34
Subject Sample	# 2	0.175	03:35
Breath Volume		1.296	03:35
Air Blank		0.000	03:35
Air Blank		0.000	03:37
Subject Sample	# 3	0.162	03:38
Breath Volume		1.304	03:38
Air Blank		0.000	03:38
Control Test		0.078	03:39
Air Blank		0.000	03:39
Diagnostics Check		OK	03:39

Data Download Date/Time: 08/07/2017 14:59

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOOHER,JOY,S
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: WOFFORD,CLINT
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 07/07/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: RAMDEO

DATE/TIME OF TEST: 07/18/2017 03:55:03

Diagnostics Check		OK	03:48
Air Blank		0.000	03:48
Control Test		0.078	03:49
Air Blank		0.000	03:49
Subject Sample	# 1	0.119	03:50
Breath Volume		1.562	03:50
Air Blank		0.000	03:51
Air Blank		0.000	03:53
Subject Sample	# 2	0.119	03:53
Breath Volume		1.445	03:53
Air Blank		0.000	03:54
Control Test		0.079	03:54
Air Blank		0.000	03:54
Diagnostics Check		OK	03:55

Data Download Date/Time: 08/07/2017 14:59

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUCAS,NATHANIEL,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: LUCAS,NATHANIEL
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 07/07/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: WEEKS

DATE/TIME OF TEST: 07/19/2017 02:51:33

Diagnostics Check		OK	02:45
Air Blank		0.000	02:45
Control Test		0.078	02:46
Air Blank		0.000	02:46
Subject Sample	# 1	0.169	02:46
Breath Volume		1.714	02:46
Air Blank		0.000	02:47
Air Blank		0.000	02:49
Subject Sample	# 2	0.177	02:50
Breath Volume		1.292	02:50
Air Blank		0.000	02:50
Control Test		0.078	02:50
Air Blank		0.000	02:51
Diagnostics Check		OK	02:51

Data Download Date/Time: 08/07/2017 14:59

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHRISTELLO,CLIFFORD,A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: CHRISTELLO,CLIFFORD
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 07/07/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: GAINER

DATE/TIME OF TEST: 07/19/2017 03:33:25

Diagnostics Check	OK	03:31	
Air Blank	0.000	03:32	
Control Test	0.078	03:32	
Air Blank		03:32	RFI Detect
Air Blank	0.000	03:33	

Data Download Date/Time: 08/07/2017 14:59

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHRISTELLO,CLIFFORD,A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: CHRISTELLO,CLIFFORD
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 07/07/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: GAINER

DATE/TIME OF TEST: 07/19/2017 03:43:15

Diagnostics Check		OK	03:35
Air Blank		0.000	03:36
Control Test		0.078	03:36
Air Blank		0.000	03:37
Subject Sample	# 1	0.103	03:38
Breath Volume		1.710	03:38
Air Blank		0.000	03:38
Air Blank		0.000	03:40
Subject Sample	# 2	0.101	03:41
Breath Volume		1.441	03:41
Air Blank		0.000	03:42
Control Test		0.078	03:42
Air Blank		0.000	03:43
Diagnostics Check		OK	03:43

Data Download Date/Time: 08/07/2017 14:59

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HEWETT,MICHAEL,W
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: HEWETT,MICHAEL
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 07/07/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 15:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:35

SUBJECT'S LAST NAME: RAKESHAW

DATE/TIME OF TEST: 07/19/2017 17:08:00

Diagnostics Check		OK	17:01
Air Blank		0.000	17:01
Control Test		0.077	17:01
Air Blank		0.000	17:02
Subject Sample	# 1	0.173	17:03
Breath Volume		1.539	17:03
Air Blank		0.000	17:04
Air Blank		0.000	17:05
Subject Sample	# 2	0.171	17:06
Breath Volume		1.636	17:06
Air Blank		0.000	17:07
Control Test		0.077	17:07
Air Blank		0.000	17:07
Diagnostics Check		OK	17:08

Data Download Date/Time: 08/07/2017 14:59

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COSTALES,ALEXANDRO,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: COSTALES,ALEXANDRO
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/07/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: HOWARD

DATE/TIME OF TEST: 07/20/2017 03:29:51

Diagnostics Check		OK	03:23
Air Blank		0.000	03:24
Control Test		0.078	03:24
Air Blank		0.000	03:24
Subject Sample	# 1	0.217	03:25
Breath Volume		2.273	03:25
Air Blank		0.000	03:25
Air Blank		0.000	03:27
Subject Sample	# 2	0.214	03:28
Breath Volume		2.332	03:28
Air Blank		0.000	03:28
Control Test		0.078	03:29
Air Blank		0.000	03:29
Diagnostics Check		OK	03:29

Data Download Date/Time: 08/07/2017 14:59

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHRISTELLO,CLIFFORD,A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: WOFFORD,CLINT
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 07/07/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:22

SUBJECT'S LAST NAME: LOPEZ-ORTIZ

DATE/TIME OF TEST: 07/21/2017 23:01:57

Diagnostics Check		OK	22:48	
Air Blank		0.000	22:49	
Control Test		0.078	22:49	
Air Blank		0.000	22:50	
Subject Sample	# 1	0.155	22:53	Volume Not Met
Breath Volume		0.156	22:53	
Air Blank		0.000	22:53	
Air Blank		0.000	22:55	
Subject Sample	# 2	0.149	22:57	
Breath Volume		2.621	22:57	
Air Blank		0.000	22:57	
Air Blank		0.000	22:59	
Subject Sample	# 3	0.145	23:00	
Breath Volume		1.757	23:00	
Air Blank		0.000	23:01	
Control Test		0.075	23:01	
Air Blank		0.000	23:01	
Diagnostics Check		OK	23:01	

Data Download Date/Time: 08/07/2017 14:59

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JEUDY,RINALDO,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: JAMES,MICHAEL
ARRESTING OFFICER AGENCY: Davenport PD
LAST AGENCY INSPECTION: 07/07/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: RODRIGUEZ-MORALES

DATE/TIME OF TEST: 07/22/2017 04:37:33

Diagnostics Check		OK	04:27
Air Blank		0.000	04:28
Control Test		0.078	04:28
Air Blank		0.000	04:29
Subject Sample	# 1	0.062	04:31
Breath Volume		1.832	04:31
Air Blank		0.000	04:32
Air Blank		0.000	04:33
Subject Sample	# 2	0.063	04:36
Breath Volume		1.910	04:36
Air Blank		0.000	04:36
Control Test		0.078	04:36
Air Blank		0.000	04:37
Diagnostics Check		OK	04:37

Data Download Date/Time: 08/07/2017 14:59

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUABA,JASON,T
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: GUABA,JASON
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/07/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:33

SUBJECT'S LAST NAME: MC MORROW

DATE/TIME OF TEST: 07/23/2017 00:01:33

Diagnostics Check		OK	23:53
Air Blank		0.000	23:54
Control Test		0.078	23:54
Air Blank		0.000	23:55
Subject Sample	# 2	0.161	23:57
Breath Volume		2.562	23:57
Air Blank		0.000	23:57
Air Blank		0.000	23:59
Subject Sample	# 1	0.160	00:00
Breath Volume		1.863	00:00
Air Blank		0.000	00:00
Control Test		0.076	00:00
Air Blank		0.000	00:01
Diagnostics Check		OK	00:01

Data Download Date/Time: 08/07/2017 14:59

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHRISTELLO,CLIFFORD,A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: CHRISTELLO,CLIFFORD
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 07/07/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: MORGAN

DATE/TIME OF TEST: 07/23/2017 22:23:53

Diagnostics Check		OK	22:20	
Air Blank		0.000	22:21	
Control Test		0.078	22:21	
Air Blank		0.000	22:21	
Subject Sample	# 1		22:23	RFI Detect
Breath Volume		0.000	22:23	
Air Blank			22:23	RFI Detect
Air Blank		0.000	22:23	

Data Download Date/Time: 08/07/2017 14:59

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHRISTELLO,CLIFFORD,A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: CHRISTELLO,CLIFFORD
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 07/07/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: MORGAN

DATE/TIME OF TEST: 07/23/2017 22:36:37

Diagnostics Check		OK	22:25	
Air Blank		0.000	22:25	
Control Test		0.078	22:25	
Air Blank		0.000	22:26	
Subject Sample	# 1		22:29	No Sample Provided
Breath Volume		0.000	22:29	
Air Blank		0.000	22:30	
Air Blank		0.000	22:32	
Subject Sample	# 2		22:35	No Sample Provided
Breath Volume		0.000	22:35	
Air Blank		0.000	22:35	
Control Test		0.079	22:36	
Air Blank		0.000	22:36	
Diagnostics Check		OK	22:36	

Data Download Date/Time: 08/07/2017 14:59

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARREN,BILLY,J
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: WARREN,BILLY
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 07/07/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:34

SUBJECT'S LAST NAME: MORRIS

DATE/TIME OF TEST: 07/24/2017 23:16:10

Diagnostics Check		OK	23:04	
Air Blank		0.000	23:05	
Control Test		0.078	23:05	
Air Blank		0.000	23:06	
Subject Sample	# 1		23:09	No Sample Provided
Breath Volume		0.000	23:09	
Air Blank		0.000	23:09	
Air Blank		0.000	23:11	
Subject Sample	# 2		23:14	No Sample Provided
Breath Volume		0.000	23:14	
Air Blank		0.000	23:15	
Control Test		0.078	23:15	
Air Blank		0.000	23:16	
Diagnostics Check		OK	23:16	

Data Download Date/Time: 08/07/2017 14:59

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUCAS,NATHANIEL,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: LUCAS,NATHANIEL
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 07/07/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 18:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:08

SUBJECT'S LAST NAME: PETERSON

DATE/TIME OF TEST: 07/27/2017 19:44:01

Diagnostics Check		OK	19:35
Air Blank		0.000	19:36
Control Test		0.079	19:36
Air Blank		0.000	19:37
Subject Sample	# 1	0.000	19:39
Breath Volume		1.523	19:39
Air Blank		0.000	19:39
Air Blank		0.000	19:41
Subject Sample	# 2	0.000	19:42
Breath Volume		1.457	19:42
Air Blank		0.000	19:43
Control Test		0.079	19:43
Air Blank		0.000	19:43
Diagnostics Check		OK	19:44

Data Download Date/Time: 08/07/2017 14:59

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COKE,CLEVELAND,A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: GEE,WESTON
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/07/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:40

SUBJECT'S LAST NAME: PUTNALL

DATE/TIME OF TEST: 07/29/2017 05:14:13

Diagnostics Check		OK	05:07
Air Blank		0.000	05:07
Control Test		0.078	05:08
Air Blank		0.000	05:08
Subject Sample	# 1	0.314	05:09
Breath Volume		1.527	05:09
Air Blank		0.000	05:10
Air Blank		0.000	05:11
Subject Sample	# 2	0.325	05:12
Breath Volume		1.781	05:12
Air Blank		0.000	05:13
Control Test		0.076	05:13
Air Blank		0.000	05:14
Diagnostics Check		OK	05:14

Data Download Date/Time: 08/07/2017 14:59

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OGG,DAVID,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: LOVETT,DOUGLAS
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 07/07/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:21

SUBJECT'S LAST NAME: BLANDIN

DATE/TIME OF TEST: 07/29/2017 06:52:04

Diagnostics Check		OK	06:47	
Air Blank		0.000	06:48	
Control Test		0.079	06:48	
Air Blank		0.000	06:49	
Subject Sample	# 1		06:51	RFI Detect
Breath Volume		0.000	06:51	
Air Blank		0.000	06:51	
			06:51	RFI Detect
Air Blank		0.000	06:52	

Data Download Date/Time: 08/07/2017 14:59

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OGG,DAVID,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: LOVETT,DOUGLAS
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 07/07/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:21

SUBJECT'S LAST NAME: BLANDIN

DATE/TIME OF TEST: 07/29/2017 07:02:17

Diagnostics Check		OK	06:54
Air Blank		0.000	06:54
Control Test		0.078	06:55
Air Blank		0.000	06:55
Subject Sample	# 1	0.108	06:57
Breath Volume		1.250	06:57
Air Blank		0.000	06:58
Air Blank		0.000	07:00
Subject Sample	# 2	0.110	07:00
Breath Volume		1.437	07:00
Air Blank		0.000	07:01
Control Test		0.078	07:01
Air Blank		0.000	07:02
Diagnostics Check		OK	07:02

Data Download Date/Time: 08/07/2017 14:59

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SYKES,STEVEN,R
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: SYKES,STEVEN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 07/07/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:03

SUBJECT'S LAST NAME: STEADMAN

DATE/TIME OF TEST: 07/31/2017 22:40:23

Diagnostics Check		OK	22:33
Air Blank		0.000	22:34
Control Test		0.077	22:34
Air Blank		0.000	22:35
Subject Sample	# 1	0.300	22:35
Breath Volume		1.179	22:35
Air Blank		0.000	22:36
Air Blank		0.000	22:38
Subject Sample	# 2	0.283	22:38
Breath Volume		1.125	22:38
Air Blank		0.000	22:39
Control Test		0.077	22:39
Air Blank		0.000	22:40
Diagnostics Check		OK	22:40

Data Download Date/Time: 08/07/2017 14:59

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001057

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUACKENBUSH,BRET,C
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: QUACKENBUSH,BRET
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 06/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: FLORES ESPINOZA

DATE/TIME OF TEST: 07/02/2017 03:34:56

Diagnostics Check		OK	03:28
Air Blank		0.000	03:29
Control Test		0.079	03:29
Air Blank		0.000	03:29
Subject Sample	# 1	0.182	03:30
Breath Volume		1.199	03:30
Air Blank		0.000	03:30
Air Blank		0.000	03:32
Subject Sample	# 2	0.188	03:33
Breath Volume		1.660	03:33
Air Blank		0.000	03:34
Control Test		0.078	03:34
Air Blank		0.000	03:34
Diagnostics Check		OK	03:34

Data Download Date/Time: 07/24/2017 10:26

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001057

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUACKENBUSH,BRET,C
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: QUACKENBUSH,BRET
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 06/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 07/03/2017 03:28:09

Diagnostics Check		OK	03:16	
Air Blank		0.000	03:17	
Control Test		0.079	03:17	
Air Blank		0.000	03:18	
Subject Sample	# 1		03:21	No Sample Provided
Breath Volume		0.000	03:21	
Air Blank		0.000	03:21	
Air Blank		0.000	03:23	
Subject Sample	# 2		03:26	No Sample Provided
Breath Volume		0.000	03:26	
Air Blank		0.000	03:27	
Control Test		0.078	03:27	
Air Blank		0.000	03:28	
Diagnostics Check		OK	03:28	

Data Download Date/Time: 07/24/2017 10:26

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001057

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTLE,CORY,A
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: SUTTLE,CORY
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 06/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: JEETOO

DATE/TIME OF TEST: 07/08/2017 00:50:16

Diagnostics Check		OK	00:40
Air Blank		0.000	00:40
Control Test		0.078	00:41
Air Blank		0.000	00:41
Subject Sample	# 1	0.195	00:44
Breath Volume		1.406	00:44
Air Blank		0.000	00:45
Air Blank		0.000	00:46
Subject Sample	# 2	0.192	00:48
Breath Volume		1.410	00:48
Air Blank		0.000	00:49
Control Test		0.077	00:49
Air Blank		0.000	00:50
Diagnostics Check		OK	00:50

Data Download Date/Time: 07/24/2017 10:26

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001057

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTLE,CORY,A
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: SUTTLE,CORY
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 06/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:38

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 07/10/2017 02:05:33

Diagnostics Check		OK	01:58
Air Blank		0.000	01:58
Control Test		0.078	01:59
Air Blank		0.000	01:59
Subject Sample	# 1	0.000	02:00
Breath Volume		1.726	02:00
Air Blank		0.000	02:00
Air Blank		0.000	02:02
Subject Sample	# 2	0.000	02:04
Breath Volume		1.714	02:04
Air Blank		0.000	02:04
Control Test		0.079	02:04
Air Blank		0.000	02:05
Diagnostics Check		OK	02:05

Data Download Date/Time: 07/24/2017 10:26

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001057

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STAFFORD,BRIAN,A
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: STAFFORD,BRIAN
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 06/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 05:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:40

SUBJECT'S LAST NAME: TURNER

DATE/TIME OF TEST: 07/11/2017 06:32:32

Diagnostics Check		OK	06:25
Air Blank		0.000	06:26
Control Test		0.077	06:26
Air Blank		0.000	06:27
Subject Sample	# 1	0.171	06:27
Breath Volume		1.914	06:27
Air Blank		0.000	06:28
Air Blank		0.000	06:30
Subject Sample	# 2	0.168	06:31
Breath Volume		1.468	06:31
Air Blank		0.000	06:31
Control Test		0.078	06:31
Air Blank		0.000	06:32
Diagnostics Check		OK	06:32

Data Download Date/Time: 07/24/2017 10:26

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001057

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STAFFORD,BRIAN,A
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: STAFFORD,BRIAN
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 06/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 05:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:40

SUBJECT'S LAST NAME: TURNER

DATE/TIME OF TEST: 07/11/2017 06:36:43

Diagnostics Check		OK	06:34	
Air Blank		0.000	06:35	
Control Test		0.078	06:35	
Air Blank		0.000	06:35	
Subject Sample	# 1		06:36	Sequence Aborted
Breath Volume		0.000	06:36	
Air Blank		0.000	06:36	

Data Download Date/Time: 07/24/2017 10:26

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001057

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTLE,CORY,A
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: SUTTLE,CORY
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 06/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: SEEPERSAD

DATE/TIME OF TEST: 07/14/2017 00:31:16

Diagnostics Check		OK	00:25
Air Blank		0.000	00:25
Control Test		0.079	00:25
Air Blank		0.000	00:26
Subject Sample	# 1	0.217	00:26
Breath Volume		1.910	00:26
Air Blank		0.000	00:27
Air Blank		0.000	00:29
Subject Sample	# 2	0.205	00:29
Breath Volume		1.871	00:29
Air Blank		0.000	00:30
Control Test		0.078	00:30
Air Blank		0.000	00:31
Diagnostics Check		OK	00:31

Data Download Date/Time: 07/24/2017 10:26

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001057

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTLE,CORY,A
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: SUTTLE,CORY
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 06/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 14:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:05

SUBJECT'S LAST NAME: BOWLING

DATE/TIME OF TEST: 07/20/2017 16:32:15

Diagnostics Check		OK	16:26
Air Blank		0.000	16:26
Control Test		0.078	16:27
Air Blank		0.000	16:27
Subject Sample	# 1	0.168	16:27
Breath Volume		2.734	16:27
Air Blank		0.000	16:28
Air Blank		0.000	16:30
Subject Sample	# 2	0.177	16:30
Breath Volume		2.589	16:30
Air Blank		0.000	16:31
Control Test		0.078	16:31
Air Blank		0.000	16:32
Diagnostics Check		OK	16:32

Data Download Date/Time: 07/24/2017 10:26

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001057

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SULLIVAN,MELANIE,A
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: SULLIVAN,MELANIE
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 06/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:04

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 07/21/2017 03:28:49

Diagnostics Check	OK	03:27	
Air Blank	0.000	03:27	
Control Test	0.078	03:28	
Air Blank		03:28	RFI Detect
Air Blank	0.000	03:28	

Data Download Date/Time: 07/24/2017 10:26

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001057

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SULLIVAN,MELANIE,A
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: SULLIVAN,MELANIE
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 06/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:04

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 07/21/2017 03:37:59

Diagnostics Check		OK	03:31
Air Blank		0.000	03:32
Control Test		0.077	03:32
Air Blank		0.000	03:32
Subject Sample	# 1	0.167	03:33
Breath Volume		1.871	03:33
Air Blank		0.000	03:34
Air Blank		0.000	03:35
Subject Sample	# 2	0.158	03:36
Breath Volume		1.792	03:36
Air Blank		0.000	03:37
Control Test		0.078	03:37
Air Blank		0.000	03:37
Diagnostics Check		OK	03:37

Data Download Date/Time: 07/24/2017 10:26

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001057

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTLE,CORY,A
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: SUTTLE,CORY
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 06/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: EDWARDS

DATE/TIME OF TEST: 07/22/2017 04:28:07

Diagnostics Check		OK	04:21
Air Blank		0.000	04:21
Control Test		0.079	04:22
Air Blank		0.000	04:22
Subject Sample	# 1	0.259	04:23
Breath Volume		1.523	04:23
Air Blank		0.000	04:23
Air Blank		0.000	04:25
Subject Sample	# 2	0.254	04:26
Breath Volume		2.203	04:26
Air Blank		0.000	04:27
Control Test		0.078	04:27
Air Blank		0.000	04:28
Diagnostics Check		OK	04:28

Data Download Date/Time: 07/24/2017 10:26

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001057

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTLE,CORY,A
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: SUTTLE,CORY
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 06/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 07/23/2017 03:26:43

Diagnostics Check		OK	03:20
Air Blank		0.000	03:21
Control Test		0.079	03:21
Air Blank		0.000	03:21
Subject Sample	# 1	0.136	03:22
Breath Volume		3.285	03:22
Air Blank		0.000	03:22
Air Blank		0.000	03:24
Subject Sample	# 2	0.140	03:25
Breath Volume		3.554	03:25
Air Blank		0.000	03:25
Control Test		0.079	03:26
Air Blank		0.000	03:26
Diagnostics Check		OK	03:26

Data Download Date/Time: 07/24/2017 10:26

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001057

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTLE,CORY,A
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: SUTTLE,CORY
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 06/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: MARTIN

DATE/TIME OF TEST: 07/24/2017 01:25:38

Diagnostics Check		OK	01:22	
Air Blank		0.000	01:23	
Control Test		0.079	01:23	
Air Blank		0.000	01:23	
Subject Sample	# 1		01:24	Subject Test Refused
Breath Volume		0.000	01:24	
Air Blank		0.000	01:24	
Control Test		0.079	01:25	
Air Blank		0.000	01:25	
Diagnostics Check		OK	01:25	

Data Download Date/Time: 07/24/2017 10:26

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001060

INSTRUMENT REGISTERED TO: WINTER HAVEN PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORPHEE,MICHAEL,
BREATH TEST OPERATOR AGENCY: Winter Haven PD
ARRESTING OFFICER: ORPHEE,MICHAEL
ARRESTING OFFICER AGENCY: Winter Haven PD
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:06

SUBJECT'S LAST NAME: CORTEZ-AVELLANEDA

DATE/TIME OF TEST: 07/01/2017 02:37:21

Diagnostics Check		OK	02:29
Air Blank		0.000	02:29
Control Test		0.081	02:30
Air Blank		0.000	02:30
Subject Sample	# 1	0.189	02:31
Breath Volume		2.011	02:31
Air Blank		0.000	02:32
Air Blank		0.000	02:34
Subject Sample	# 2	0.183	02:35
Breath Volume		1.187	02:35
Air Blank		0.000	02:36
Control Test		0.084	02:36
Air Blank		0.000	02:37
Diagnostics Check		OK	02:37

Data Download Date/Time: 07/25/2017 16:02

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001060

INSTRUMENT REGISTERED TO: WINTER HAVEN PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORPHEE,MICHAEL,
BREATH TEST OPERATOR AGENCY: Winter Haven PD
ARRESTING OFFICER: CRUZ,EDGARDO
ARRESTING OFFICER AGENCY: Winter Haven PD
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:28

SUBJECT'S LAST NAME: FRIEDLANDER

DATE/TIME OF TEST: 07/07/2017 03:09:28

Diagnostics Check		OK	03:00
Air Blank		0.000	03:01
Control Test		0.082	03:01
Air Blank		0.000	03:02
Subject Sample	# 1	0.147	03:03
Breath Volume		2.300	03:03
Air Blank		0.000	03:04
Air Blank		0.000	03:05
Subject Sample	# 2	0.140	03:07
Breath Volume		3.007	03:07
Air Blank		0.000	03:08
Control Test		0.082	03:08
Air Blank		0.000	03:09
Diagnostics Check		OK	03:09

Data Download Date/Time: 07/25/2017 16:02

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001060

INSTRUMENT REGISTERED TO: WINTER HAVEN PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORPHEE,MICHAEL,
BREATH TEST OPERATOR AGENCY: Winter Haven PD
ARRESTING OFFICER: ORPHEE,MICHAEL
ARRESTING OFFICER AGENCY: Winter Haven PD
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:11

SUBJECT'S LAST NAME: BRYANT

DATE/TIME OF TEST: 07/15/2017 04:47:08

Diagnostics Check		OK	04:35	
Air Blank		0.000	04:36	
Control Test		0.082	04:36	
Air Blank		0.000	04:36	
Subject Sample	# 1	0.113	04:40	Volume Not Met
Breath Volume		0.570	04:40	
Air Blank		0.000	04:40	
Air Blank		0.000	04:42	
Subject Sample	# 2		04:45	No Sample Provided
Breath Volume		0.000	04:45	
Air Blank		0.000	04:46	
Control Test		0.080	04:46	
Air Blank		0.000	04:47	
Diagnostics Check		OK	04:47	

Data Download Date/Time: 07/25/2017 16:02

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001060

INSTRUMENT REGISTERED TO: WINTER HAVEN PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORPHEE,MICHAEL,
BREATH TEST OPERATOR AGENCY: Winter Haven PD
ARRESTING OFFICER: ORPHEE,MICHAEL
ARRESTING OFFICER AGENCY: Winter Haven PD
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:11

SUBJECT'S LAST NAME: BRYANT

DATE/TIME OF TEST: 07/15/2017 05:05:54

Diagnostics Check		OK	04:48	
Air Blank		0.000	04:49	
Control Test		0.080	04:49	
Air Blank		0.000	04:50	
Subject Sample	# 1	0.131	04:52	
Breath Volume		2.132	04:52	
Air Blank		0.000	04:53	
Air Blank		0.000	04:55	
Subject Sample	# 2	0.133	04:58	Volume Not Met
Breath Volume		0.507	04:58	
Air Blank		0.000	04:59	
Air Blank		0.000	05:00	
Subject Sample	# 3	0.141	05:04	Volume Not Met
Breath Volume		0.257	05:04	
Air Blank		0.000	05:04	
Control Test		0.081	05:05	
Air Blank		0.000	05:05	
Diagnostics Check		OK	05:05	

Data Download Date/Time: 07/25/2017 16:02

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001060

INSTRUMENT REGISTERED TO: WINTER HAVEN PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BONSALL,ROBERT,E
BREATH TEST OPERATOR AGENCY: Winter Haven PD
ARRESTING OFFICER: BOONE,BRANDON
ARRESTING OFFICER AGENCY: Winter Haven PD
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 07/22/2017 00:28:12

Diagnostics Check		OK	00:25	
Air Blank		0.000	00:25	
Control Test		0.081	00:25	
Air Blank		0.000	00:26	
Subject Sample	# 1	0.191	00:26	
Breath Volume		3.578	00:26	
Air Blank			00:27	Purge Fail
Air Blank		0.000	00:28	

Data Download Date/Time: 07/25/2017 16:02

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001060

INSTRUMENT REGISTERED TO: WINTER HAVEN PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BONSALL,ROBERT,E
BREATH TEST OPERATOR AGENCY: Winter Haven PD
ARRESTING OFFICER: BOONE,BRANDON
ARRESTING OFFICER AGENCY: Winter Haven PD
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 07/22/2017 00:38:19

Diagnostics Check		OK	00:31
Air Blank		0.000	00:31
Control Test		0.081	00:32
Air Blank		0.000	00:32
Subject Sample	# 1	0.185	00:33
Breath Volume		3.519	00:33
Air Blank		0.000	00:33
Air Blank		0.000	00:35
Subject Sample	# 2	0.194	00:36
Breath Volume		4.796	00:36
Air Blank		0.000	00:37
Control Test		0.083	00:37
Air Blank		0.000	00:38
Diagnostics Check		OK	00:38

Data Download Date/Time: 07/25/2017 16:02

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001060

INSTRUMENT REGISTERED TO: WINTER HAVEN PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BONSALL,ROBERT,E
BREATH TEST OPERATOR AGENCY: Winter Haven PD
ARRESTING OFFICER: BOONE,BRANDON
ARRESTING OFFICER AGENCY: Winter Haven PD
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:16

SUBJECT'S LAST NAME: HILL

DATE/TIME OF TEST: 07/23/2017 22:47:44

Diagnostics Check		OK	22:41
Air Blank		0.000	22:41
Control Test		0.082	22:41
Air Blank		0.000	22:42
Subject Sample	# 1	0.151	22:43
Breath Volume		2.570	22:43
Air Blank		0.000	22:43
Air Blank		0.000	22:45
Subject Sample	# 2	0.133	22:46
Breath Volume		2.855	22:46
Air Blank		0.000	22:46
Control Test		0.082	22:47
Air Blank		0.000	22:47
Diagnostics Check		OK	22:47

Data Download Date/Time: 07/25/2017 16:02

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001061

INSTRUMENT REGISTERED TO: GULFPORT PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAROTTA,MICHAEL,S
BREATH TEST OPERATOR AGENCY: Gulfport PD
ARRESTING OFFICER: MAROTTA,MICHAEL
ARRESTING OFFICER AGENCY: Gulfport PD
LAST AGENCY INSPECTION: 06/07/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A3
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:41

SUBJECT'S LAST NAME: VAZQUEZ

DATE/TIME OF TEST: 07/01/2017 01:08:32

Diagnostics Check		OK	01:02
Air Blank		0.000	01:02
Control Test		0.080	01:02
Air Blank		0.000	01:03
Subject Sample	# 1	0.130	01:04
Breath Volume		2.136	01:04
Air Blank		0.000	01:04
Air Blank		0.000	01:06
Subject Sample	# 2	0.128	01:07
Breath Volume		1.847	01:07
Air Blank		0.000	01:07
Control Test		0.079	01:07
Air Blank		0.000	01:08
Diagnostics Check		OK	01:08

Data Download Date/Time: 07/21/2017 22:13

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001062

INSTRUMENT REGISTERED TO: TREASURE ISLAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUDTKE,CHRISTOPHER,E
BREATH TEST OPERATOR AGENCY: OTHER State
ARRESTING OFFICER: LUDTKE,CHRISTOPHER
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 06/19/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 12716080A6
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: BUI
ARREST TIME: 13:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:21

SUBJECT'S LAST NAME: CATTERTON

DATE/TIME OF TEST: 07/04/2017 14:53:28

Diagnostics Check		OK	14:50	
Air Blank		0.000	14:51	
Control Test		0.079	14:51	
Air Blank		0.000	14:51	
Subject Sample	# 1		14:52	Subject Test Refused
Breath Volume		0.000	14:52	
Air Blank		0.000	14:52	
Control Test		0.080	14:52	
Air Blank		0.000	14:53	
Diagnostics Check		OK	14:53	

Data Download Date/Time: 07/21/2017 09:37

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001062

INSTRUMENT REGISTERED TO: TREASURE ISLAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALVIS,REECE,H
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)
ARRESTING OFFICER: ALVIS,REECE
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 06/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A6
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: BUI
ARREST TIME: 18:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:12

SUBJECT'S LAST NAME: SCHAFFER

DATE/TIME OF TEST: 07/04/2017 19:43:53

Diagnostics Check		OK	19:37
Air Blank		0.000	19:38
Control Test		0.080	19:38
Air Blank		0.000	19:38
Subject Sample	# 1	0.060	19:39
Breath Volume		2.707	19:39
Air Blank		0.000	19:40
Air Blank		0.000	19:41
Subject Sample	# 2	0.057	19:42
Breath Volume		1.566	19:42
Air Blank		0.000	19:42
Control Test		0.080	19:43
Air Blank		0.000	19:43
Diagnostics Check		OK	19:43

Data Download Date/Time: 07/21/2017 09:37

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001062

INSTRUMENT REGISTERED TO: TREASURE ISLAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIAZ,HECTOR,
BREATH TEST OPERATOR AGENCY: Treasure Island PD
ARRESTING OFFICER: DIAZ,HECTOR
ARRESTING OFFICER AGENCY: Treasure Island PD
LAST AGENCY INSPECTION: 06/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A6
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:00

SUBJECT'S LAST NAME: MURRAY

DATE/TIME OF TEST: 07/09/2017 05:26:45

Diagnostics Check		OK	05:23	
Air Blank		0.000	05:24	
Control Test		0.080	05:24	
Air Blank		0.000	05:25	
Subject Sample	# 1		05:25	Subject Test Refused
Breath Volume		0.000	05:25	
Air Blank		0.000	05:25	
Control Test		0.080	05:26	
Air Blank		0.000	05:26	
Diagnostics Check		OK	05:26	

Data Download Date/Time: 07/21/2017 09:37

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001063

INSTRUMENT REGISTERED TO: HOLLYWOOD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DERIET,RAFAEL,
BREATH TEST OPERATOR AGENCY: Hollywood PD
ARRESTING OFFICER: DERIET,RAFAEL
ARRESTING OFFICER AGENCY: Hollywood PD
LAST AGENCY INSPECTION: 06/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: ALLEN

DATE/TIME OF TEST: 07/22/2017 03:09:45

Diagnostics Check		OK	03:03
Air Blank		0.000	03:03
Control Test		0.077	03:04
Air Blank		0.000	03:04
Subject Sample	# 1	0.190	03:05
Breath Volume		3.750	03:05
Air Blank		0.000	03:05
Air Blank		0.000	03:07
Subject Sample	# 2	0.183	03:08
Breath Volume		4.066	03:08
Air Blank		0.000	03:08
Control Test		0.077	03:09
Air Blank		0.000	03:09
Diagnostics Check		OK	03:09

Data Download Date/Time: 07/31/2017 10:46

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

INSTRUMENT REGISTERED TO: NEW PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,ROBERT,R
BREATH TEST OPERATOR AGENCY: Port Richey PD
ARRESTING OFFICER: BERG,TIMOTHY
ARRESTING OFFICER AGENCY: New Port Richey PD
LAST AGENCY INSPECTION: 06/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:48

SUBJECT'S LAST NAME: CHIPELO

DATE/TIME OF TEST: 07/06/2017 05:17:34

Diagnostics Check		OK	05:08
Air Blank		0.000	05:09
Control Test		0.079	05:09
Air Blank		0.000	05:10
Subject Sample	# 1	0.152	05:13
Breath Volume		2.718	05:13
Air Blank		0.000	05:13
Air Blank		0.000	05:15
Subject Sample	# 2	0.157	05:16
Breath Volume		2.128	05:16
Air Blank		0.000	05:16
Control Test		0.078	05:16
Air Blank		0.000	05:17
Diagnostics Check		OK	05:17

Data Download Date/Time: 07/28/2017 00:16

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

INSTRUMENT REGISTERED TO: NEW PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRAPNELL,CHRIS,P
BREATH TEST OPERATOR AGENCY: New Port Richey PD
ARRESTING OFFICER: GREEN,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:13

SUBJECT'S LAST NAME: NORTON

DATE/TIME OF TEST: 07/18/2017 23:58:45

Diagnostics Check		OK	23:52
Air Blank		0.000	23:52
Control Test		0.078	23:53
Air Blank		0.000	23:53
Subject Sample	# 1	0.193	23:54
Breath Volume		1.738	23:54
Air Blank		0.000	23:54
Air Blank		0.000	23:56
Subject Sample	# 2	0.184	23:57
Breath Volume		1.675	23:57
Air Blank		0.000	23:57
Control Test		0.078	23:58
Air Blank		0.000	23:58
Diagnostics Check		OK	23:58

Data Download Date/Time: 07/28/2017 00:16

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001065

INSTRUMENT REGISTERED TO: TARPON SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEMMON,CHRISTOPHER,F
BREATH TEST OPERATOR AGENCY: Tarpon Springs PD
ARRESTING OFFICER: LEMMON,CHRISTOPHER
ARRESTING OFFICER AGENCY: Tarpon Springs PD
LAST AGENCY INSPECTION: 06/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A3
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: MOO

DATE/TIME OF TEST: 07/09/2017 04:17:35

Diagnostics Check		OK	04:06	
Air Blank		0.000	04:06	
Control Test		0.079	04:06	
Air Blank		0.000	04:07	
Subject Sample	# 1	0.181	04:10	Volume Not Met
Breath Volume		0.179	04:10	
Air Blank		0.000	04:11	
Air Blank		0.000	04:12	
Subject Sample	# 2	0.143	04:16	Volume Not Met
Breath Volume		0.027	04:16	
Air Blank		0.000	04:16	
Control Test		0.078	04:16	
Air Blank		0.000	04:17	
Diagnostics Check		OK	04:17	

Data Download Date/Time: 07/24/2017 13:49

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001065

INSTRUMENT REGISTERED TO: TARPON SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEMMON,CHRISTOPHER,F
BREATH TEST OPERATOR AGENCY: Tarpon Springs PD
ARRESTING OFFICER: LEMMON,CHRISTOPHER
ARRESTING OFFICER AGENCY: Tarpon Springs PD
LAST AGENCY INSPECTION: 06/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A3
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: MOO

DATE/TIME OF TEST: 07/09/2017 04:32:56

Diagnostics Check		OK	04:21	
Air Blank		0.000	04:21	
Control Test		0.078	04:22	
Air Blank		0.000	04:22	
Subject Sample	# 1	0.182	04:25	Volume Not Met
Breath Volume		0.117	04:25	
Air Blank		0.000	04:26	
Air Blank		0.000	04:28	
Subject Sample	# 2	0.145	04:31	Volume Not Met
Breath Volume		0.625	04:31	
Air Blank		0.000	04:31	
Control Test		0.078	04:32	
Air Blank		0.000	04:32	
Diagnostics Check		OK	04:32	

Data Download Date/Time: 07/24/2017 13:49

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001065

INSTRUMENT REGISTERED TO: TARPON SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIBSON,JOHN,
BREATH TEST OPERATOR AGENCY: Tarpon Springs PD
ARRESTING OFFICER: RONDOS,ROBERT
ARRESTING OFFICER AGENCY: Tarpon Springs PD
LAST AGENCY INSPECTION: 06/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A3
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 16:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:33

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 07/12/2017 17:08:02

Diagnostics Check		OK	16:53
Air Blank		0.000	16:54
Control Test		0.079	16:54
Air Blank		0.000	16:55
Subject Sample	# 1	0.282	16:56
Breath Volume		1.531	16:56
Air Blank		0.000	16:57
Air Blank		0.000	16:58
Subject Sample	# 2	0.260	17:01
Breath Volume		2.015	17:01
Air Blank		0.000	17:01
Air Blank		0.000	17:03
Subject Sample	# 3	0.270	17:06
Breath Volume		1.722	17:06
Air Blank		0.000	17:07
Control Test		0.077	17:07
Air Blank		0.000	17:07
Diagnostics Check		OK	17:08

Data Download Date/Time: 07/24/2017 13:49

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001065

INSTRUMENT REGISTERED TO: TARPON SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPATZ,JOHN,M
BREATH TEST OPERATOR AGENCY: Tarpon Springs PD
ARRESTING OFFICER: SPATZ,JOHN
ARRESTING OFFICER AGENCY: Tarpon Springs PD
LAST AGENCY INSPECTION: 07/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A3
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: OLIVARES

DATE/TIME OF TEST: 07/26/2017 03:09:52

Diagnostics Check		OK	03:07	
Air Blank		0.000	03:07	
Control Test		0.080	03:07	
Air Blank		0.000	03:08	
Subject Sample	# 1		03:08	Subject Test Refused
Breath Volume		0.000	03:08	
Air Blank		0.000	03:08	
Control Test		0.080	03:09	
Air Blank		0.000	03:09	
Diagnostics Check		OK	03:09	

Data Download Date/Time: 08/28/2017 13:53

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WHITEHURST,W
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 22:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:05

SUBJECT'S LAST NAME: LUNN

DATE/TIME OF TEST: 07/26/2017 23:32:20

Diagnostics Check		OK	23:25
Air Blank		0.000	23:26
Control Test		0.080	23:26
Air Blank		0.000	23:27
Subject Sample	# 1	0.168	23:27
Breath Volume		1.632	23:27
Air Blank		0.000	23:28
Air Blank		0.000	23:30
Subject Sample	# 2	0.167	23:30
Breath Volume		1.890	23:30
Air Blank		0.000	23:31
Control Test		0.080	23:31
Air Blank		0.000	23:32
Diagnostics Check		OK	23:32

Data Download Date/Time: 07/28/2017 20:34

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NUGENT,DAVID,A
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: NUGENT,DAVID
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 06/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: TOWNSEND

DATE/TIME OF TEST: 07/01/2017 02:01:01

Diagnostics Check		OK	01:53
Air Blank		0.000	01:54
Control Test		0.080	01:54
Air Blank		0.000	01:55
Subject Sample	# 1	0.105	01:55
Breath Volume		1.312	01:55
Air Blank		0.000	01:56
Air Blank		0.000	01:58
Subject Sample	# 2	0.107	01:59
Breath Volume		1.480	01:59
Air Blank		0.000	02:00
Control Test		0.080	02:00
Air Blank		0.000	02:00
Diagnostics Check		OK	02:01

Data Download Date/Time: 07/25/2017 00:33

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NUGENT,DAVID,A
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: NUGENT,DAVID
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 06/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:16

SUBJECT'S LAST NAME: CLENDENIN

DATE/TIME OF TEST: 07/03/2017 01:58:20

Diagnostics Check		OK	01:53	
Air Blank		0.000	01:53	
Control Test		0.079	01:54	
Air Blank		0.000	01:54	
Subject Sample	# 1		01:56	Subject Test Refused
Breath Volume		0.000	01:56	
Air Blank		0.000	01:57	
Control Test		0.080	01:57	
Air Blank		0.000	01:58	
Diagnostics Check		OK	01:58	

Data Download Date/Time: 07/25/2017 00:33

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NEGERSMITH,D,M
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: NEGERSMITH,D
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 06/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:07

SUBJECT'S LAST NAME: BLAISDELL

DATE/TIME OF TEST: 07/08/2017 23:48:32

Diagnostics Check		OK	23:41
Air Blank		0.000	23:41
Control Test		0.080	23:42
Air Blank		0.000	23:42
Subject Sample	# 1	0.000	23:43
Breath Volume		2.031	23:43
Air Blank		0.000	23:44
Air Blank		0.000	23:46
Subject Sample	# 2	0.000	23:47
Breath Volume		2.214	23:47
Air Blank		0.000	23:47
Control Test		0.080	23:47
Air Blank		0.000	23:48
Diagnostics Check		OK	23:48

Data Download Date/Time: 07/25/2017 00:33

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,JARROD,J
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: YOUNG,JARROD
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 06/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:06

SUBJECT'S LAST NAME: CARVER

DATE/TIME OF TEST: 07/13/2017 04:33:44

Diagnostics Check		OK	04:27
Air Blank		0.000	04:27
Control Test		0.079	04:28
Air Blank		0.000	04:28
Subject Sample	# 1	0.142	04:29
Breath Volume		2.253	04:29
Air Blank		0.000	04:30
Air Blank		0.000	04:31
Subject Sample	# 2	0.145	04:32
Breath Volume		2.832	04:32
Air Blank		0.000	04:32
Control Test		0.079	04:33
Air Blank		0.000	04:33
Diagnostics Check		OK	04:33

Data Download Date/Time: 07/25/2017 00:33

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRUNEAU,OFC,
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: BRUNEAU,OFC
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 06/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: SHAWHAN

DATE/TIME OF TEST: 07/14/2017 04:42:39

Diagnostics Check		OK	04:35
Air Blank		0.000	04:36
Control Test		0.080	04:36
Air Blank		0.000	04:36
Subject Sample	# 1	0.182	04:37
Breath Volume		3.984	04:37
Air Blank		0.000	04:38
Air Blank		0.000	04:40
Subject Sample	# 2	0.183	04:40
Breath Volume		4.226	04:40
Air Blank		0.000	04:41
Control Test		0.079	04:42
Air Blank		0.000	04:42
Diagnostics Check		OK	04:42

Data Download Date/Time: 07/25/2017 00:33

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ENGLAND,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: ENGLAND,CHRISTOPHER
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 06/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 18:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:07

SUBJECT'S LAST NAME: GRIFFIN

DATE/TIME OF TEST: 07/15/2017 19:33:34

Diagnostics Check		OK	19:30	
Air Blank		0.000	19:30	
Control Test		0.080	19:30	
Air Blank		0.000	19:31	
Subject Sample	# 1		19:32	Subject Test Refused
Breath Volume		0.000	19:32	
Air Blank		0.000	19:32	
Control Test		0.080	19:32	
Air Blank		0.000	19:33	
Diagnostics Check		OK	19:33	

Data Download Date/Time: 07/25/2017 00:33

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ENGLAND,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: ENGLAND,CHRISTOPHER
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 06/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:30

SUBJECT'S LAST NAME: WASHBURN

DATE/TIME OF TEST: 07/19/2017 20:02:49

Diagnostics Check		OK	19:51	
Air Blank		0.000	19:51	
Control Test		0.080	19:51	
Air Blank		0.000	19:52	
Subject Sample	# 1	0.105	19:55	Volume Not Met
Breath Volume		0.234	19:55	
Air Blank		0.000	19:56	
Air Blank		0.000	19:57	
Subject Sample	# 2	0.087	20:01	Volume Not Met
Breath Volume		0.562	20:01	
Air Blank		0.000	20:01	
Control Test		0.079	20:02	
Air Blank		0.000	20:02	
Diagnostics Check		OK	20:02	

Data Download Date/Time: 07/25/2017 00:33

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ENGLAND,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: ENGLAND,CHRISTOPHER
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 06/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:30

SUBJECT'S LAST NAME: WASHBURN

DATE/TIME OF TEST: 07/19/2017 20:11:54

Diagnostics Check		OK	20:04
Air Blank		0.000	20:04
Control Test		0.078	20:05
Air Blank		0.000	20:05
Subject Sample	# 1	0.095	20:06
Breath Volume		1.511	20:06
Air Blank		0.000	20:07
Air Blank		0.000	20:09
Subject Sample	# 2	0.093	20:09
Breath Volume		1.976	20:09
Air Blank		0.000	20:10
Control Test		0.078	20:11
Air Blank		0.000	20:11
Diagnostics Check		OK	20:11

Data Download Date/Time: 07/25/2017 00:33

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ENGLAND,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: ENGLAND,CHRISTOPHER
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 06/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: PACHECO

DATE/TIME OF TEST: 07/22/2017 01:01:09

Diagnostics Check		OK	00:51
Air Blank		0.000	00:51
Control Test		0.080	00:51
Air Blank		0.000	00:52
Subject Sample	# 1	0.134	00:53
Breath Volume		1.914	00:53
Air Blank		0.000	00:54
Air Blank		0.000	00:55
Subject Sample	# 2	0.112	00:56
Breath Volume		1.332	00:56
Air Blank		0.000	00:57
Air Blank		0.000	00:58
Subject Sample	# 3	0.126	00:59
Breath Volume		2.015	00:59
Air Blank		0.000	01:00
Control Test		0.080	01:00
Air Blank		0.000	01:01
Diagnostics Check		OK	01:01

Data Download Date/Time: 07/25/2017 00:33

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NEGERSMITH,D,M
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: NEGERSMITH,D
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 06/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: COUTURE

DATE/TIME OF TEST: 07/22/2017 23:19:17

Diagnostics Check		OK	23:12
Air Blank		0.000	23:12
Control Test		0.080	23:13
Air Blank		0.000	23:13
Subject Sample	# 1	0.000	23:14
Breath Volume		2.621	23:14
Air Blank		0.000	23:14
Air Blank		0.000	23:16
Subject Sample	# 2	0.000	23:17
Breath Volume		2.406	23:17
Air Blank		0.000	23:18
Control Test		0.079	23:18
Air Blank		0.000	23:19
Diagnostics Check		OK	23:19

Data Download Date/Time: 07/25/2017 00:34

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NUGENT,DAVE,
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: NUGENT,DAVE
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 06/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: HICKMAN

DATE/TIME OF TEST: 07/23/2017 04:07:53

Diagnostics Check		OK	04:04	
Air Blank		0.000	04:05	
Control Test		0.080	04:05	
Air Blank		0.000	04:06	
Subject Sample	# 1		04:06	Subject Test Refused
Breath Volume		0.000	04:06	
Air Blank		0.000	04:06	
Control Test		0.080	04:07	
Air Blank		0.000	04:07	
Diagnostics Check		OK	04:07	

Data Download Date/Time: 07/25/2017 00:34

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NUGENT,DAVID,A
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: HISA,N
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 06/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:40

SUBJECT'S LAST NAME: HOLCOMB

DATE/TIME OF TEST: 07/23/2017 05:07:39

Diagnostics Check		OK	05:00
Air Blank		0.000	05:01
Control Test		0.079	05:01
Air Blank		0.000	05:02
Subject Sample	# 1	0.137	05:02
Breath Volume		1.105	05:02
Air Blank		0.000	05:03
Air Blank		0.000	05:05
Subject Sample	# 2	0.131	05:06
Breath Volume		1.570	05:06
Air Blank		0.000	05:06
Control Test		0.079	05:07
Air Blank		0.000	05:07
Diagnostics Check		OK	05:07

Data Download Date/Time: 07/25/2017 00:34

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NEGERSMITH,D,M
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: NEGERSMITH,D
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 06/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:51

SUBJECT'S LAST NAME: ACEVEDO QUINONES

DATE/TIME OF TEST: 07/23/2017 05:26:13

Diagnostics Check		OK	05:20	
Air Blank		0.000	05:21	
Control Test		0.079	05:21	
Air Blank		0.000	05:22	
Subject Sample	# 1	0.170	05:24	
Breath Volume		1.656	05:24	
Air Blank			05:25	Purge Fail
Air Blank		0.000	05:26	

Data Download Date/Time: 07/25/2017 00:34

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NEGERSMITH,D,M
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: NEGERSMITH,D
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 06/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:51

SUBJECT'S LAST NAME: ACEVEDO QUINONES

DATE/TIME OF TEST: 07/23/2017 05:39:00

Diagnostics Check		OK	05:28
Air Blank		0.000	05:28
Control Test		0.076	05:29
Air Blank		0.000	05:29
Subject Sample	# 1	0.176	05:31
Breath Volume		2.722	05:31
Air Blank		0.000	05:32
Air Blank		0.000	05:34
Subject Sample	# 2	0.158	05:37
Breath Volume		2.046	05:37
Air Blank		0.000	05:38
Control Test		0.079	05:38
Air Blank		0.000	05:38
Diagnostics Check		OK	05:39

Data Download Date/Time: 07/25/2017 00:34

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001071

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ASHE,KEVIN,B
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: PRITCHARD,T
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 06/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: LABAW

DATE/TIME OF TEST: 07/01/2017 01:38:53

Diagnostics Check		OK	01:31
Air Blank		0.000	01:31
Control Test		0.080	01:31
Air Blank		0.000	01:32
Subject Sample	# 1	0.149	01:32
Breath Volume		1.937	01:32
Air Blank		0.000	01:33
Air Blank		0.000	01:35
Subject Sample	# 2	0.147	01:37
Breath Volume		1.144	01:37
Air Blank		0.000	01:37
Control Test		0.080	01:38
Air Blank		0.000	01:38
Diagnostics Check		OK	01:38

Data Download Date/Time: 07/23/2017 16:48

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001071

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ASHE,KEVIN,B
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: MCREE,M
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 06/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:27

SUBJECT'S LAST NAME: MINARD

DATE/TIME OF TEST: 07/01/2017 04:56:52

Diagnostics Check		OK	04:50
Air Blank		0.000	04:50
Control Test		0.079	04:51
Air Blank		0.000	04:51
Subject Sample	# 1	0.198	04:52
Breath Volume		1.746	04:52
Air Blank		0.000	04:52
Air Blank		0.000	04:54
Subject Sample	# 2	0.191	04:55
Breath Volume		1.339	04:55
Air Blank		0.000	04:55
Control Test		0.078	04:56
Air Blank		0.000	04:56
Diagnostics Check		OK	04:56

Data Download Date/Time: 07/23/2017 16:48

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001071

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRAUNLICH,G,A
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: MILLER,G
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 06/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 17:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:26

SUBJECT'S LAST NAME: HAYES

DATE/TIME OF TEST: 07/10/2017 18:53:05

Diagnostics Check		OK	18:46
Air Blank		0.000	18:46
Control Test		0.079	18:47
Air Blank		0.000	18:47
Subject Sample	# 1	0.000	18:48
Breath Volume		2.167	18:48
Air Blank		0.000	18:48
Air Blank		0.000	18:50
Subject Sample	# 2	0.000	18:51
Breath Volume		1.476	18:51
Air Blank		0.000	18:52
Control Test		0.079	18:52
Air Blank		0.000	18:53
Diagnostics Check		OK	18:53

Data Download Date/Time: 07/23/2017 16:48

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001071

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ASHE,KEVIN,B
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: MILLER,C
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 06/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: ORTIZ

DATE/TIME OF TEST: 07/17/2017 03:17:27

Diagnostics Check		OK	03:10
Air Blank		0.000	03:11
Control Test		0.079	03:11
Air Blank		0.000	03:12
Subject Sample	# 1	0.184	03:12
Breath Volume		3.136	03:12
Air Blank		0.000	03:13
Air Blank		0.000	03:15
Subject Sample	# 2	0.180	03:15
Breath Volume		3.015	03:15
Air Blank		0.000	03:16
Control Test		0.080	03:16
Air Blank		0.000	03:17
Diagnostics Check		OK	03:17

Data Download Date/Time: 07/23/2017 16:48

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001071

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALTSTATT,MATTHEW,B
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: SAUNDERS,J
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 06/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:59

SUBJECT'S LAST NAME: SHIEBLER

DATE/TIME OF TEST: 07/18/2017 00:27:19

Diagnostics Check		OK	00:20
Air Blank		0.000	00:21
Control Test		0.079	00:21
Air Blank		0.000	00:21
Subject Sample	# 1	0.156	00:22
Breath Volume		2.226	00:22
Air Blank		0.000	00:23
Air Blank		0.000	00:25
Subject Sample	# 2	0.165	00:25
Breath Volume		2.035	00:25
Air Blank		0.000	00:26
Control Test		0.079	00:26
Air Blank		0.000	00:27
Diagnostics Check		OK	00:27

Data Download Date/Time: 07/23/2017 16:48

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001071

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALTSTATT,MATTHEW,B
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: SYDOROWICZ,J
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 06/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:08

SUBJECT'S LAST NAME: CORBIN

DATE/TIME OF TEST: 07/19/2017 01:34:48

Diagnostics Check		OK	01:28
Air Blank		0.000	01:28
Control Test		0.080	01:29
Air Blank		0.000	01:29
Subject Sample	# 1	0.138	01:30
Breath Volume		2.253	01:30
Air Blank		0.000	01:30
Air Blank		0.000	01:32
Subject Sample	# 2	0.133	01:33
Breath Volume		1.890	01:33
Air Blank		0.000	01:33
Control Test		0.079	01:34
Air Blank		0.000	01:34
Diagnostics Check		OK	01:34

Data Download Date/Time: 07/23/2017 16:48

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001071

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINSON,C,A
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: SAUNDERS,J
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 06/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 17:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:40

SUBJECT'S LAST NAME: STARLING

DATE/TIME OF TEST: 07/22/2017 18:17:33

Diagnostics Check		OK	18:11
Air Blank		0.000	18:11
Control Test		0.079	18:11
Air Blank		0.000	18:12
Subject Sample	# 1	0.170	18:13
Breath Volume		1.742	18:13
Air Blank		0.000	18:13
Air Blank		0.000	18:15
Subject Sample	# 2	0.159	18:16
Breath Volume		3.218	18:16
Air Blank		0.000	18:16
Control Test		0.079	18:16
Air Blank		0.000	18:17
Diagnostics Check		OK	18:17

Data Download Date/Time: 07/23/2017 16:48

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001071

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALTSTATT,MATTHEW,B
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: SYDOROWICZ,J
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 06/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:51

SUBJECT'S LAST NAME: MEJIA

DATE/TIME OF TEST: 07/23/2017 04:17:48

Diagnostics Check		OK	04:11
Air Blank		0.000	04:11
Control Test		0.079	04:12
Air Blank		0.000	04:12
Subject Sample	# 1	0.210	04:13
Breath Volume		2.152	04:13
Air Blank		0.000	04:13
Air Blank		0.000	04:15
Subject Sample	# 2	0.198	04:16
Breath Volume		2.429	04:16
Air Blank		0.000	04:16
Control Test		0.079	04:17
Air Blank		0.000	04:17
Diagnostics Check		OK	04:17

Data Download Date/Time: 07/23/2017 16:48

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001071

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRAUNLICH,G,A
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: SOLIS,E
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 07/23/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:01

SUBJECT'S LAST NAME: BERNARD

DATE/TIME OF TEST: 07/26/2017 03:27:56

Diagnostics Check		OK	03:21
Air Blank		0.000	03:22
Control Test		0.080	03:22
Air Blank		0.000	03:23
Subject Sample	# 1	0.063	03:23
Breath Volume		3.167	03:23
Air Blank		0.000	03:24
Air Blank		0.000	03:25
Subject Sample	# 2	0.055	03:26
Breath Volume		1.292	03:26
Air Blank		0.000	03:27
Control Test		0.079	03:27
Air Blank		0.000	03:27
Diagnostics Check		OK	03:27

Data Download Date/Time: 08/22/2017 17:22

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001071

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALTSTATT,MATTHEW,B
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: BRASWELL,J
ARRESTING OFFICER AGENCY: Green Cove Springs
LAST AGENCY INSPECTION: 07/23/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 19:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:30

SUBJECT'S LAST NAME: MULLINS

DATE/TIME OF TEST: 07/27/2017 20:04:12

Diagnostics Check		OK	19:57
Air Blank		0.000	19:57
Control Test		0.080	19:57
Air Blank		0.000	19:58
Subject Sample	# 1	0.000	19:59
Breath Volume		2.226	19:59
Air Blank		0.000	19:59
Air Blank		0.000	20:02
Subject Sample	# 2	0.000	20:02
Breath Volume		2.531	20:02
Air Blank		0.000	20:03
Control Test		0.080	20:03
Air Blank		0.000	20:04
Diagnostics Check		OK	20:04

Data Download Date/Time: 08/22/2017 17:22

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001074

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STONE,CHRIS,R
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: FRANCIS,DEP.
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 06/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A2
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 05:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:57

SUBJECT'S LAST NAME: NEAL

DATE/TIME OF TEST: 07/08/2017 07:26:33

Diagnostics Check		OK	07:17
Air Blank		0.000	07:17
Control Test		0.078	07:18
Air Blank		0.000	07:18
Subject Sample	# 1	0.139	07:19
Breath Volume		2.000	07:19
Air Blank		0.000	07:19
Air Blank		0.000	07:21
Subject Sample	# 2	0.170	07:22
Breath Volume		2.890	07:22
Air Blank		0.000	07:22
Air Blank		0.000	07:24
Subject Sample	# 3	0.165	07:24
Breath Volume		2.324	07:24
Air Blank		0.000	07:25
Control Test		0.077	07:25
Air Blank		0.000	07:26
Diagnostics Check		OK	07:26

Data Download Date/Time: 07/26/2017 12:12

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001074

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRIOLAT,CASEY,J
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: BRIOLAT,CASEY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A2
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 05:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:28

SUBJECT'S LAST NAME: VALENZUELA CUBA

DATE/TIME OF TEST: 07/30/2017 06:56:52

Diagnostics Check		OK	06:49
Air Blank		0.000	06:49
Control Test		0.077	06:50
Air Blank		0.000	06:50
Subject Sample	# 1	0.138	06:52
Breath Volume		2.574	06:52
Air Blank		0.000	06:52
Air Blank		0.000	06:54
Subject Sample	# 2	0.127	06:55
Breath Volume		2.535	06:55
Air Blank		0.000	06:55
Control Test		0.077	06:56
Air Blank		0.000	06:56
Diagnostics Check		OK	06:56

Data Download Date/Time: 08/23/2017 13:30

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001075

INSTRUMENT REGISTERED TO: FHP TROOP E
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORBETA,JORGE,
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: ORBETA,JORGE
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 06/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1824329
DRY GAS EXPIRY DATE: 03/09/2018
VIOLATION CODE: DUI
ARREST TIME: 00:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:24

SUBJECT'S LAST NAME: FERRO

DATE/TIME OF TEST: 07/01/2017 01:06:58

Diagnostics Check		OK	01:00
Air Blank		0.000	01:00
Control Test		0.079	01:01
Air Blank		0.000	01:01
Subject Sample	# 1	0.175	01:02
Breath Volume		3.496	01:02
Air Blank		0.000	01:03
Air Blank		0.000	01:04
Subject Sample	# 2	0.169	01:05
Breath Volume		3.511	01:05
Air Blank		0.000	01:06
Control Test		0.078	01:06
Air Blank		0.000	01:06
Diagnostics Check		OK	01:06

Data Download Date/Time: 07/07/2017 01:19

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001075

INSTRUMENT REGISTERED TO: FHP TROOP E
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORBETA,JORGE,
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: ORBETA,JORGE
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 06/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1824329
DRY GAS EXPIRY DATE: 03/09/2018
VIOLATION CODE: DUI
ARREST TIME: 21:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:36

SUBJECT'S LAST NAME: SCHILD

DATE/TIME OF TEST: 07/03/2017 22:10:56

Diagnostics Check		OK	22:05	
Air Blank		0.000	22:06	
Control Test		0.079	22:06	
Air Blank		0.000	22:07	
Subject Sample	# 1		22:09	Subject Test Refused
Breath Volume		0.000	22:09	
Air Blank		0.000	22:09	
Control Test		0.079	22:10	
Air Blank		0.000	22:10	
Diagnostics Check		OK	22:10	

Data Download Date/Time: 07/07/2017 01:19

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001075

INSTRUMENT REGISTERED TO: FHP TROOP E
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,SHAWN,W
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: MARTIN,SHAWN
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 06/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1824329
DRY GAS EXPIRY DATE: 03/09/2018
VIOLATION CODE: DUI
ARREST TIME: 15:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:02

SUBJECT'S LAST NAME: MONTGOMERY

DATE/TIME OF TEST: 07/04/2017 15:26:13

Diagnostics Check		OK	15:23	
Air Blank		0.000	15:23	
Control Test		0.079	15:24	
Air Blank		0.000	15:24	
Subject Sample	# 1		15:24	Subject Test Refused
Breath Volume		0.000	15:24	
Air Blank		0.000	15:25	
Control Test		0.079	15:25	
Air Blank		0.000	15:26	
Diagnostics Check		OK	15:26	

Data Download Date/Time: 07/07/2017 01:19

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001075

INSTRUMENT REGISTERED TO: FHP TROOP E
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORBETA,JORGE,
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: ORBETA,JORGE
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 06/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1824329
DRY GAS EXPIRY DATE: 03/09/2018
VIOLATION CODE: DUI
ARREST TIME: 19:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:18

SUBJECT'S LAST NAME: WATERMAN

DATE/TIME OF TEST: 07/04/2017 19:46:27

Diagnostics Check		OK	19:38
Air Blank		0.000	19:39
Control Test		0.078	19:39
Air Blank		0.000	19:40
Subject Sample	# 1	0.026	19:41
Breath Volume		1.726	19:41
Air Blank		0.000	19:41
Air Blank		0.000	19:43
Subject Sample	# 2	0.029	19:44
Breath Volume		2.105	19:44
Air Blank		0.000	19:45
Control Test		0.079	19:45
Air Blank		0.000	19:46
Diagnostics Check		OK	19:46

Data Download Date/Time: 07/07/2017 01:19

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001076

INSTRUMENT REGISTERED TO: ZEPHYRHILLS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARDNER,NATHAN,R
BREATH TEST OPERATOR AGENCY: Zephyrhills PD
ARRESTING OFFICER: CARMER,WILLIAM
ARRESTING OFFICER AGENCY: Zephyrhills PD
LAST AGENCY INSPECTION: 06/09/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: WEST

DATE/TIME OF TEST: 07/15/2017 22:35:36

Diagnostics Check		OK	22:29
Air Blank		0.000	22:29
Control Test		0.079	22:30
Air Blank		0.000	22:30
Subject Sample	# 1	0.224	22:31
Breath Volume		1.593	22:31
Air Blank		0.000	22:31
Air Blank		0.000	22:33
Subject Sample	# 2	0.207	22:34
Breath Volume		1.386	22:34
Air Blank		0.000	22:34
Control Test		0.078	22:35
Air Blank		0.000	22:35
Diagnostics Check		OK	22:35

Data Download Date/Time: 07/29/2017 03:37

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001077

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASSICK,PATRICIA,A
BREATH TEST OPERATOR AGENCY: Hardee County SO
ARRESTING OFFICER: SPENCER,ROBERT
ARRESTING OFFICER AGENCY: Wauchula PD
LAST AGENCY INSPECTION: 06/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 00:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:17

SUBJECT'S LAST NAME: LY

DATE/TIME OF TEST: 07/02/2017 00:58:39

Diagnostics Check		OK	00:52
Air Blank		0.000	00:53
Control Test		0.080	00:53
Air Blank		0.000	00:53
Subject Sample	# 1	0.182	00:54
Breath Volume		1.976	00:54
Air Blank		0.000	00:54
Air Blank		0.000	00:56
Subject Sample	# 2	0.184	00:57
Breath Volume		1.921	00:57
Air Blank		0.000	00:57
Control Test		0.080	00:58
Air Blank		0.000	00:58
Diagnostics Check		OK	00:58

Data Download Date/Time: 07/26/2017 12:01

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZOLLOTA,VLADIMIR,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: KATAYIANES,MICHAEL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 07/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:09

SUBJECT'S LAST NAME: RISMOND

DATE/TIME OF TEST: 07/07/2017 21:42:50

Diagnostics Check		OK	21:36
Air Blank		0.000	21:37
Control Test		0.080	21:37
Air Blank		0.000	21:38
Subject Sample	# 1	0.148	21:38
Breath Volume		1.488	21:38
Air Blank		0.000	21:39
Air Blank		0.000	21:40
Subject Sample	# 2	0.151	21:41
Breath Volume		1.917	21:41
Air Blank		0.000	21:41
Control Test		0.079	21:42
Air Blank		0.000	21:42
Diagnostics Check		OK	21:42

Data Download Date/Time: 08/05/2017 17:24

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLI,ANTHONY,N
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: KARAYIANES,MICHAEL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 07/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: MCCLENDON

DATE/TIME OF TEST: 07/14/2017 00:19:01

Diagnostics Check		OK	00:07
Air Blank		0.000	00:07
Control Test		0.079	00:08
Air Blank		0.000	00:08
Subject Sample	# 1	0.181	00:09
Breath Volume		1.972	00:09
Air Blank		0.000	00:10
Air Blank		0.000	00:11
Subject Sample	# 2	0.154	00:13
Breath Volume		1.851	00:13
Air Blank		0.000	00:14
Air Blank		0.000	00:16
Subject Sample	# 3	0.161	00:17
Breath Volume		1.742	00:17
Air Blank		0.000	00:17
Control Test		0.078	00:18
Air Blank		0.000	00:18
Diagnostics Check		OK	00:19

Data Download Date/Time: 08/05/2017 17:24

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZOLLOTA,VLADIMIR,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ALLI,ANTHONY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 07/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:32

SUBJECT'S LAST NAME: PENNYWELL

DATE/TIME OF TEST: 07/14/2017 03:08:24

Diagnostics Check		OK	03:02
Air Blank		0.000	03:02
Control Test		0.079	03:02
Air Blank		0.000	03:03
Subject Sample	# 1	0.160	03:03
Breath Volume		1.601	03:03
Air Blank		0.000	03:04
Air Blank		0.000	03:06
Subject Sample	# 2	0.155	03:06
Breath Volume		1.562	03:06
Air Blank		0.000	03:07
Control Test		0.080	03:07
Air Blank		0.000	03:08
Diagnostics Check		OK	03:08

Data Download Date/Time: 08/05/2017 17:24

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COX,LENARD,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ZOLLOTA,VLADIMIR
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 07/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:18

SUBJECT'S LAST NAME: KOROSEC

DATE/TIME OF TEST: 07/14/2017 19:39:12

Diagnostics Check	OK	19:38	
Air Blank		19:38	RFI Detect
Air Blank	0.000	19:39	

Data Download Date/Time: 08/05/2017 17:24

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COX,LENARD,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ZOLLOTA,VLADIMIR
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 07/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:18

SUBJECT'S LAST NAME: KOROSEC

DATE/TIME OF TEST: 07/14/2017 19:54:46

Diagnostics Check		OK	19:40	
Air Blank		0.000	19:40	
Control Test		0.080	19:41	
Air Blank		0.000	19:41	
Subject Sample	# 1	0.172	19:44	Volume Not Met
Breath Volume		0.628	19:44	
Air Blank		0.000	19:45	
Air Blank		0.000	19:47	
Subject Sample	# 2	0.194	19:49	
Breath Volume		1.234	19:49	
Air Blank		0.000	19:50	
Air Blank		0.000	19:51	
Subject Sample	# 3	0.204	19:53	
Breath Volume		1.312	19:53	
Air Blank		0.000	19:53	
Control Test		0.078	19:54	
Air Blank		0.000	19:54	
Diagnostics Check		OK	19:54	

Data Download Date/Time: 08/05/2017 17:24

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZOLLOTA,VLADIMIR,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: COX,LENARD
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 07/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: BENNETT

DATE/TIME OF TEST: 07/15/2017 02:34:40

Diagnostics Check		OK	02:31	
Air Blank		0.000	02:31	
Control Test		0.082	02:32	
Air Blank		0.000	02:32	
Subject Sample	# 1		02:33	Subject Test Refused
Breath Volume		0.000	02:33	
Air Blank		0.000	02:33	
Control Test		0.080	02:34	
Air Blank		0.000	02:34	
Diagnostics Check		OK	02:34	

Data Download Date/Time: 08/05/2017 17:24

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZOLLOTA,VLADIMIR,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ALLI,ANTHONY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 07/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:24

SUBJECT'S LAST NAME: BADIE

DATE/TIME OF TEST: 07/15/2017 03:09:14

Diagnostics Check		OK	03:01
Air Blank		0.000	03:02
Control Test		0.080	03:02
Air Blank		0.000	03:03
Subject Sample	# 1	0.176	03:03
Breath Volume		1.457	03:03
Air Blank		0.000	03:03
Air Blank		0.000	03:05
Subject Sample	# 2	0.184	03:07
Breath Volume		1.199	03:07
Air Blank		0.000	03:08
Control Test		0.080	03:08
Air Blank		0.000	03:09
Diagnostics Check		OK	03:09

Data Download Date/Time: 08/05/2017 17:24

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZOLLOTA,VLADIMIR,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: WEISKOPF,MICHAEL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 07/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: CARVALLO

DATE/TIME OF TEST: 07/16/2017 01:11:36

Diagnostics Check		OK	01:05
Air Blank		0.000	01:05
Control Test		0.080	01:06
Air Blank		0.000	01:06
Subject Sample	# 1	0.161	01:06
Breath Volume		1.492	01:06
Air Blank		0.000	01:07
Air Blank		0.000	01:09
Subject Sample	# 2	0.166	01:10
Breath Volume		1.957	01:10
Air Blank		0.000	01:10
Control Test		0.079	01:11
Air Blank		0.000	01:11
Diagnostics Check		OK	01:11

Data Download Date/Time: 08/05/2017 17:24

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COX,LENARD,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ZOLLOTA,VLADIMIR
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 07/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: CRENSHAW

DATE/TIME OF TEST: 07/16/2017 02:57:07

Diagnostics Check	OK	02:55	
Air Blank	0.000	02:56	
Control Test	0.080	02:56	RFI Detect
Air Blank	0.000	02:57	

Data Download Date/Time: 08/05/2017 17:24

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COX,LENARD,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ZOLLOTA,VLADIMIR
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 07/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: CRENSHAW

DATE/TIME OF TEST: 07/16/2017 03:05:15

Diagnostics Check		OK	02:58
Air Blank		0.000	02:58
Control Test		0.079	02:59
Air Blank		0.000	02:59
Subject Sample	# 1	0.198	03:00
Breath Volume		2.792	03:00
Air Blank		0.000	03:01
Air Blank		0.000	03:03
Subject Sample	# 2	0.179	03:03
Breath Volume		1.105	03:03
Air Blank		0.000	03:04
Control Test		0.079	03:04
Air Blank		0.000	03:05
Diagnostics Check		OK	03:05

Data Download Date/Time: 08/05/2017 17:24

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEISKOPF,MICHAEL,D
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ALLI,ANTHONY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 07/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:18

SUBJECT'S LAST NAME: DESTEFANO

DATE/TIME OF TEST: 07/27/2017 22:46:08

Diagnostics Check		OK	22:40
Air Blank		0.000	22:40
Control Test		0.079	22:40
Air Blank		0.000	22:41
Subject Sample	# 1	0.132	22:41
Breath Volume		1.386	22:41
Air Blank		0.000	22:42
Air Blank		0.000	22:44
Subject Sample	# 2	0.137	22:44
Breath Volume		1.156	22:44
Air Blank		0.000	22:45
Control Test		0.078	22:45
Air Blank		0.000	22:46
Diagnostics Check		OK	22:46

Data Download Date/Time: 08/05/2017 17:24

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLI,ANTHONY,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: WEISKOPF,MICHAEL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 07/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:13

SUBJECT'S LAST NAME: COPELAND JR

DATE/TIME OF TEST: 07/29/2017 02:53:23

Diagnostics Check		OK	02:46
Air Blank		0.000	02:46
Control Test		0.080	02:47
Air Blank		0.000	02:47
Subject Sample	# 1	0.256	02:48
Breath Volume		1.167	02:48
Air Blank		0.000	02:49
Air Blank		0.000	02:51
Subject Sample	# 2	0.238	02:51
Breath Volume		1.398	02:51
Air Blank		0.000	02:52
Control Test		0.079	02:52
Air Blank		0.000	02:53
Diagnostics Check		OK	02:53

Data Download Date/Time: 08/05/2017 17:24

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEISKOPF,MICHAEL,D
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: KARAYIANES,MICHAEL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 07/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: TRUGLIO

DATE/TIME OF TEST: 07/29/2017 04:02:45

Diagnostics Check		OK	03:59	
Air Blank		0.000	03:59	
Control Test		0.080	04:00	
Air Blank		0.000	04:00	
Subject Sample	# 1		04:01	Subject Test Refused
Breath Volume		0.000	04:01	
Air Blank		0.000	04:01	
Control Test		0.079	04:02	
Air Blank		0.000	04:02	
Diagnostics Check		OK	04:02	

Data Download Date/Time: 08/05/2017 17:24

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001082

INSTRUMENT REGISTERED TO: PLANTATION PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BEAUREGARD,JEFF,
BREATH TEST OPERATOR AGENCY: Plantation PD
ARRESTING OFFICER: MCWILLIAMS,RYAN
ARRESTING OFFICER AGENCY: Plantation PD
LAST AGENCY INSPECTION: 06/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A2
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: MOORE

DATE/TIME OF TEST: 07/14/2017 23:14:46

Diagnostics Check		OK	23:11	
Air Blank		0.000	23:12	
Control Test		0.079	23:12	
Air Blank		0.000	23:13	
Subject Sample	# 1		23:13	Subject Test Refused
Breath Volume		0.000	23:13	
Air Blank		0.000	23:13	
Control Test		0.078	23:14	
Air Blank		0.000	23:14	
Diagnostics Check		OK	23:14	

Data Download Date/Time: 07/29/2017 14:26

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001109

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALFY,RACHEL,M
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: PALFY,RACHEL
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A3
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: DAIGLE

DATE/TIME OF TEST: 07/15/2017 00:51:01

Diagnostics Check		OK	00:44
Air Blank		0.000	00:44
Control Test		0.077	00:45
Air Blank		0.000	00:45
Subject Sample	# 1	0.224	00:46
Breath Volume		3.000	00:46
Air Blank		0.000	00:47
Air Blank		0.000	00:48
Subject Sample	# 2	0.209	00:49
Breath Volume		2.519	00:49
Air Blank		0.000	00:50
Control Test		0.078	00:50
Air Blank		0.000	00:50
Diagnostics Check		OK	00:51

Data Download Date/Time: 07/31/2017 18:56

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001109

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALFY,RACHEL,M
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: PALFY,RACHEL
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A3
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: WHITNEY

DATE/TIME OF TEST: 07/15/2017 23:32:47

Diagnostics Check		OK	23:26
Air Blank		0.000	23:27
Control Test		0.077	23:27
Air Blank		0.000	23:27
Subject Sample	# 1	0.172	23:28
Breath Volume		3.027	23:28
Air Blank		0.000	23:28
Air Blank		0.000	23:30
Subject Sample	# 2	0.169	23:31
Breath Volume		2.640	23:31
Air Blank		0.000	23:31
Control Test		0.077	23:32
Air Blank		0.000	23:32
Diagnostics Check		OK	23:32

Data Download Date/Time: 07/31/2017 18:56

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001109

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALFY,RACHEL,M
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: PALFY,RACHEL
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A3
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:02

SUBJECT'S LAST NAME: VEGA RIVERA

DATE/TIME OF TEST: 07/16/2017 04:32:17

Diagnostics Check		OK	04:25
Air Blank		0.000	04:26
Control Test		0.077	04:26
Air Blank		0.000	04:27
Subject Sample	# 1	0.145	04:27
Breath Volume		1.968	04:27
Air Blank		0.000	04:28
Air Blank		0.000	04:30
Subject Sample	# 2	0.136	04:30
Breath Volume		1.335	04:30
Air Blank		0.000	04:31
Control Test		0.077	04:31
Air Blank		0.000	04:32
Diagnostics Check		OK	04:32

Data Download Date/Time: 07/31/2017 18:56

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001112

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSON,CLEVELAND,L
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: DULIN,EDWARD
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 06/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 11:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 12:45

SUBJECT'S LAST NAME: ROSE

DATE/TIME OF TEST: 07/22/2017 13:24:26

Diagnostics Check		OK	13:17
Air Blank		0.000	13:17
Control Test		0.081	13:17
Air Blank		0.000	13:18
Subject Sample	# 1	0.200	13:19
Breath Volume		1.253	13:19
Air Blank		0.000	13:20
Air Blank		0.000	13:21
Subject Sample	# 2	0.219	13:22
Breath Volume		2.328	13:22
Air Blank		0.000	13:23
Control Test		0.080	13:23
Air Blank		0.000	13:24
Diagnostics Check		OK	13:24

Data Download Date/Time: 07/31/2017 13:49

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001112

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWARD,DOMINIQUE,L
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: HOWARD,DOMINIQUE
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 06/11/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 19:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: HAMPTON

DATE/TIME OF TEST: 07/24/2017 21:35:31

Diagnostics Check		OK	21:29
Air Blank		0.000	21:29
Control Test		0.081	21:30
Air Blank		0.000	21:30
Subject Sample	# 1	0.222	21:31
Breath Volume		3.687	21:31
Air Blank		0.000	21:31
Air Blank		0.000	21:33
Subject Sample	# 2	0.210	21:34
Breath Volume		3.414	21:34
Air Blank		0.000	21:34
Control Test		0.081	21:34
Air Blank		0.000	21:35
Diagnostics Check		OK	21:35

Data Download Date/Time: 07/31/2017 13:49

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001112

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HITCHCOCK,WILLIAM,D
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: HITCHCOCK,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 06/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 01:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: NISWONGER

DATE/TIME OF TEST: 07/27/2017 01:52:40

Diagnostics Check		OK	01:45
Air Blank		0.000	01:46
Control Test		0.082	01:46
Air Blank		0.000	01:47
Subject Sample	# 1	0.000	01:48
Breath Volume		1.679	01:48
Air Blank		0.000	01:48
Air Blank		0.000	01:50
Subject Sample	# 2	0.000	01:51
Breath Volume		1.457	01:51
Air Blank		0.000	01:51
Control Test		0.082	01:52
Air Blank		0.000	01:52
Diagnostics Check		OK	01:52

Data Download Date/Time: 07/31/2017 13:49

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001112

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HITCHCOCK,WILLIAM,D
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: MULLIGAN,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 06/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 22:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: FERRANTE

DATE/TIME OF TEST: 07/28/2017 23:44:04

Diagnostics Check		OK	23:38	
Air Blank		0.000	23:38	
Control Test		0.082	23:38	
Air Blank		0.000	23:39	
Subject Sample	# 1	0.000	23:40	
Breath Volume		3.718	23:40	
Air Blank		0.000	23:40	
Air Blank		0.000	23:42	
Subject Sample	# 2		23:42	Subject Test Refused
Breath Volume		0.000	23:42	
Air Blank		0.000	23:43	
Control Test		0.082	23:43	
Air Blank		0.000	23:43	
Diagnostics Check		OK	23:44	

Data Download Date/Time: 07/31/2017 13:49

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001113

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,ALAIN,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: HERNANDEZ,ALAIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:59

SUBJECT'S LAST NAME: CARDONA

DATE/TIME OF TEST: 07/01/2017 02:21:49

Diagnostics Check		OK	02:19	
Air Blank		0.000	02:20	
Control Test		0.082	02:20	
Air Blank		0.000	02:20	
Subject Sample	# 1		02:21	RFI Detect
Breath Volume		0.000	02:21	
Air Blank			02:21	RFI Detect
Air Blank		0.000	02:21	

Data Download Date/Time: 07/31/2017 19:00

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001113

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,ALAIN,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: HERNANDEZ,ALAIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:59

SUBJECT'S LAST NAME: CARDONA

DATE/TIME OF TEST: 07/01/2017 02:29:01

Diagnostics Check		OK	02:22
Air Blank		0.000	02:23
Control Test		0.082	02:23
Air Blank		0.000	02:24
Subject Sample	# 1	0.090	02:24
Breath Volume		2.082	02:24
Air Blank		0.000	02:25
Air Blank		0.000	02:26
Subject Sample	# 2	0.085	02:27
Breath Volume		1.253	02:27
Air Blank		0.000	02:28
Control Test		0.082	02:28
Air Blank		0.000	02:28
Diagnostics Check		OK	02:29

Data Download Date/Time: 07/31/2017 19:00

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001113

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALFY,RACHEL,M
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: PALFY,RACHEL
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: MELGOZA

DATE/TIME OF TEST: 07/01/2017 03:49:56

Diagnostics Check		OK	03:43
Air Blank		0.000	03:44
Control Test		0.081	03:44
Air Blank		0.000	03:44
Subject Sample	# 1	0.166	03:45
Breath Volume		2.222	03:45
Air Blank		0.000	03:46
Air Blank		0.000	03:47
Subject Sample	# 2	0.172	03:48
Breath Volume		1.828	03:48
Air Blank		0.000	03:49
Control Test		0.082	03:49
Air Blank		0.000	03:49
Diagnostics Check		OK	03:49

Data Download Date/Time: 07/31/2017 19:00

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001113

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,ALAIN,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: HERNANDEZ,ALAIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 02:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: BLANTON

DATE/TIME OF TEST: 07/02/2017 02:31:30

Diagnostics Check		OK	02:25
Air Blank		0.000	02:26
Control Test		0.081	02:26
Air Blank		0.000	02:26
Subject Sample	# 1	0.163	02:27
Breath Volume		2.898	02:27
Air Blank		0.000	02:27
Air Blank		0.000	02:29
Subject Sample	# 2	0.157	02:30
Breath Volume		2.539	02:30
Air Blank		0.000	02:30
Control Test		0.081	02:30
Air Blank		0.000	02:31
Diagnostics Check		OK	02:31

Data Download Date/Time: 07/31/2017 19:00

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001113

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,ALAIN,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: HERNANDEZ,ALAIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 03:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:43

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 07/02/2017 04:17:39

Diagnostics Check		OK	04:03	
Air Blank		0.000	04:03	
Control Test		0.081	04:04	
Air Blank		0.000	04:04	
Subject Sample	# 1	0.189	04:07	Volume Not Met
Breath Volume		0.167	04:07	
Air Blank		0.000	04:08	
Air Blank		0.000	04:10	
Subject Sample	# 2	0.207	04:12	
Breath Volume		1.910	04:12	
Air Blank		0.000	04:12	
Air Blank		0.000	04:14	
Subject Sample	# 3	0.198	04:16	
Breath Volume		1.226	04:16	
Air Blank		0.000	04:16	
Control Test		0.080	04:17	
Air Blank		0.000	04:17	
Diagnostics Check		OK	04:17	

Data Download Date/Time: 07/31/2017 19:01

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001113

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,ALAIN,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: PALFY,RACHEL
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 18:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:52

SUBJECT'S LAST NAME: BRUNSON

DATE/TIME OF TEST: 07/02/2017 18:13:16

Diagnostics Check	OK	18:12	
Air Blank		18:12	RFI Detect
Air Blank	0.000	18:13	

Data Download Date/Time: 07/31/2017 19:01

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001113

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,ALAIN,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: PALFY,RACHEL
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 18:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:52

SUBJECT'S LAST NAME: BRUNSON

DATE/TIME OF TEST: 07/02/2017 18:20:53

Diagnostics Check		OK	18:14
Air Blank		0.000	18:15
Control Test		0.082	18:15
Air Blank		0.000	18:16
Subject Sample	# 1	0.075	18:16
Breath Volume		1.753	18:16
Air Blank		0.000	18:17
Air Blank		0.000	18:19
Subject Sample	# 2	0.075	18:19
Breath Volume		2.394	18:19
Air Blank		0.000	18:19
Control Test		0.081	18:20
Air Blank		0.000	18:20
Diagnostics Check		OK	18:20

Data Download Date/Time: 07/31/2017 19:01

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001113

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,ALAIN,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: HERNANDEZ,ALAIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 18:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:44

SUBJECT'S LAST NAME: HERNANDEZ DIAZ

DATE/TIME OF TEST: 07/04/2017 19:11:04

Diagnostics Check		OK	19:04
Air Blank		0.000	19:05
Control Test		0.081	19:05
Air Blank		0.000	19:06
Subject Sample	# 1	0.299	19:06
Breath Volume		2.082	19:06
Air Blank		0.000	19:07
Air Blank		0.000	19:09
Subject Sample	# 2	0.312	19:09
Breath Volume		1.277	19:09
Air Blank		0.000	19:10
Control Test		0.081	19:10
Air Blank		0.000	19:10
Diagnostics Check		OK	19:11

Data Download Date/Time: 07/31/2017 19:01

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001113

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,ALAI,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: HERNANDEZ,ALAIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 22:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:08

SUBJECT'S LAST NAME: SEGUNDO SEGUNDO

DATE/TIME OF TEST: 07/08/2017 22:31:32

Diagnostics Check		OK	22:28	
Air Blank		0.000	22:28	
Control Test		0.081	22:29	
Air Blank		0.000	22:29	
Subject Sample	# 1		22:30	Subject Test Refused
Breath Volume		0.000	22:30	
Air Blank		0.000	22:30	
Control Test		0.082	22:30	
Air Blank		0.000	22:31	
Diagnostics Check		OK	22:31	

Data Download Date/Time: 07/31/2017 19:01

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001113

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,ALAIN,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: HERNANDEZ,ALAIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:12

SUBJECT'S LAST NAME: VARGAS-VALLEJO

DATE/TIME OF TEST: 07/09/2017 01:44:56

Diagnostics Check		OK	01:32	
Air Blank		0.000	01:33	
Control Test		0.082	01:33	
Air Blank		0.000	01:34	
Subject Sample	# 1	0.165	01:37	Volume Not Met
Breath Volume		0.398	01:37	
Air Blank		0.000	01:37	
Air Blank		0.000	01:39	
Subject Sample	# 2	0.170	01:40	
Breath Volume		1.519	01:40	
Air Blank		0.000	01:40	
Air Blank		0.000	01:42	
Subject Sample	# 3	0.180	01:43	
Breath Volume		2.046	01:43	
Air Blank		0.000	01:44	
Control Test		0.081	01:44	
Air Blank		0.000	01:44	
Diagnostics Check		OK	01:44	

Data Download Date/Time: 07/31/2017 19:01

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001113

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,ALAIN,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: HERNANDEZ,ALAIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:44

SUBJECT'S LAST NAME: CARDONA

DATE/TIME OF TEST: 07/15/2017 03:17:30

Diagnostics Check		OK	03:04	
Air Blank		0.000	03:05	
Control Test		0.082	03:05	
Air Blank		0.000	03:06	
Subject Sample	# 1	0.171	03:09	Volume Not Met
Breath Volume		0.878	03:09	
Air Blank		0.000	03:09	
Air Blank		0.000	03:11	
Subject Sample	# 2	0.213	03:12	
Breath Volume		1.667	03:12	
Air Blank		0.000	03:13	
Air Blank		0.000	03:15	
Subject Sample	# 3	0.204	03:16	
Breath Volume		2.214	03:16	
Air Blank		0.000	03:16	
Control Test		0.080	03:16	
Air Blank		0.000	03:17	
Diagnostics Check		OK	03:17	

Data Download Date/Time: 07/31/2017 19:01

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001113

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,ALAIN,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: HERNANDEZ,ALAIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 02:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:48

SUBJECT'S LAST NAME: MCGEE

DATE/TIME OF TEST: 07/16/2017 02:24:48

Diagnostics Check		OK	02:18
Air Blank		0.000	02:19
Control Test		0.082	02:19
Air Blank		0.000	02:19
Subject Sample	# 1	0.174	02:20
Breath Volume		2.300	02:20
Air Blank		0.000	02:20
Air Blank		0.000	02:22
Subject Sample	# 2	0.166	02:23
Breath Volume		2.003	02:23
Air Blank		0.000	02:23
Control Test		0.082	02:24
Air Blank		0.000	02:24
Diagnostics Check		OK	02:24

Data Download Date/Time: 07/31/2017 19:01

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001113

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,ALAIN,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: HERNANDEZ,ALAIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 03:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:54

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 07/16/2017 04:19:03

Diagnostics Check	OK	04:17	
Air Blank	0.000	04:18	
Control Test	0.081	04:18	RFI Detect
Air Blank	0.000	04:19	

Data Download Date/Time: 07/31/2017 19:01

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001113

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,ALAIN,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: HERNANDEZ,ALAIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 03:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:54

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 07/16/2017 04:24:13

Diagnostics Check		OK	04:20	
Air Blank		0.000	04:21	
Control Test		0.081	04:21	
Air Blank		0.000	04:22	
Subject Sample	# 1		04:22	Subject Test Refused
Breath Volume		0.785	04:22	
Air Blank		0.000	04:23	
Control Test		0.082	04:23	
Air Blank		0.000	04:24	
Diagnostics Check		OK	04:24	

Data Download Date/Time: 07/31/2017 19:01

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001113

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,ALAIN,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: HERNANDEZ,ALAIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:44

SUBJECT'S LAST NAME: WOLFE

DATE/TIME OF TEST: 07/23/2017 00:17:09

Diagnostics Check		OK	00:08
Air Blank		0.000	00:08
Control Test		0.081	00:08
Air Blank		0.000	00:09
Subject Sample	# 1	0.146	00:10
Breath Volume		2.742	00:10
Air Blank		0.000	00:10
Air Blank		0.000	00:12
Subject Sample	# 2	0.122	00:12
Breath Volume		2.414	00:12
Air Blank		0.000	00:13
Air Blank		0.000	00:15
Subject Sample	# 3	0.144	00:15
Breath Volume		2.078	00:15
Air Blank		0.000	00:16
Control Test		0.081	00:16
Air Blank		0.000	00:17
Diagnostics Check		OK	00:17

Data Download Date/Time: 07/31/2017 19:01

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001113

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,ALAIN,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 23:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:47

SUBJECT'S LAST NAME: PAOLINETTI

DATE/TIME OF TEST: 07/23/2017 00:19:53

Diagnostics Check	OK	00:18	
Air Blank	0.000	00:18	
Control Test	0.080	00:18	
Air Blank		00:19	RFI Detect
Air Blank	0.000	00:19	

Data Download Date/Time: 07/31/2017 19:01

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001113

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,ALAIN,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 23:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:47

SUBJECT'S LAST NAME: PAOLINETTI

DATE/TIME OF TEST: 07/23/2017 00:27:33

Diagnostics Check		OK	00:20
Air Blank		0.000	00:21
Control Test		0.081	00:21
Air Blank		0.000	00:22
Subject Sample	# 1	0.045	00:22
Breath Volume		1.410	00:22
Air Blank		0.000	00:23
Air Blank		0.000	00:25
Subject Sample	# 2	0.041	00:26
Breath Volume		1.238	00:26
Air Blank		0.000	00:26
Control Test		0.082	00:26
Air Blank		0.000	00:27
Diagnostics Check		OK	00:27

Data Download Date/Time: 07/31/2017 19:01

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001113

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,ALAIN,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: HERNANDEZ,ALAIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 02:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: MARIN AVILES

DATE/TIME OF TEST: 07/30/2017 02:46:44

Diagnostics Check		OK	02:40
Air Blank		0.000	02:41
Control Test		0.081	02:41
Air Blank		0.000	02:42
Subject Sample	# 1	0.104	02:42
Breath Volume		2.390	02:42
Air Blank		0.000	02:43
Air Blank		0.000	02:44
Subject Sample	# 2	0.100	02:45
Breath Volume		1.433	02:45
Air Blank		0.000	02:45
Control Test		0.081	02:46
Air Blank		0.000	02:46
Diagnostics Check		OK	02:46

Data Download Date/Time: 07/31/2017 19:01

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001113

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,ALAIN,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: HERNANDEZ,ALAIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 17:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:42

SUBJECT'S LAST NAME: UBALDO ALONSO

DATE/TIME OF TEST: 07/30/2017 18:07:31

Diagnostics Check		OK	18:04	
Air Blank		0.000	18:04	
Control Test		0.080	18:04	
Air Blank		0.000	18:05	
Subject Sample	# 1		18:05	RFI Detect
Breath Volume		1.121	18:05	
Air Blank		0.000	18:06	
Control Test		0.081	18:06	
Air Blank		0.000	18:07	
Diagnostics Check		OK	18:07	

Data Download Date/Time: 07/31/2017 19:01

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001113

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,ALAIN,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: HERNANDEZ,ALAIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 17:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:42

SUBJECT'S LAST NAME: UBALDO ALONSO

DATE/TIME OF TEST: 07/30/2017 18:15:05

Diagnostics Check		OK	18:08
Air Blank		0.000	18:09
Control Test		0.080	18:09
Air Blank		0.000	18:09
Subject Sample	# 1	0.243	18:10
Breath Volume		1.863	18:10
Air Blank		0.000	18:10
Air Blank		0.000	18:12
Subject Sample	# 2	0.251	18:13
Breath Volume		3.308	18:13
Air Blank		0.000	18:14
Control Test		0.082	18:14
Air Blank		0.000	18:14
Diagnostics Check		OK	18:15

Data Download Date/Time: 07/31/2017 19:01

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001114

INSTRUMENT REGISTERED TO: DELAND P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WISE,B,A
BREATH TEST OPERATOR AGENCY: Deland PD
ARRESTING OFFICER: OWENS,T
ARRESTING OFFICER AGENCY: Deland PD
LAST AGENCY INSPECTION: 07/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG616503
DRY GAS EXPIRY DATE: 06/13/2018
VIOLATION CODE: DUI
ARREST TIME: 21:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: BARKER

DATE/TIME OF TEST: 07/27/2017 22:00:38

Diagnostics Check		OK	21:57	
Air Blank		0.000	21:58	
Control Test		0.077	21:58	
Air Blank		0.000	21:59	
Subject Sample	# 1		21:59	Subject Test Refused
Breath Volume		0.000	21:59	
Air Blank		0.000	21:59	
Control Test		0.076	22:00	
Air Blank		0.000	22:00	
Diagnostics Check		OK	22:00	

Data Download Date/Time: 08/29/2017 14:13

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001114

INSTRUMENT REGISTERED TO: DELAND P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MULERO,JOHAN,
BREATH TEST OPERATOR AGENCY: Deland PD
ARRESTING OFFICER: SALMON,MATTHEW
ARRESTING OFFICER AGENCY: Deland PD
LAST AGENCY INSPECTION: 07/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG616503
DRY GAS EXPIRY DATE: 06/13/2018
VIOLATION CODE: DUI
ARREST TIME: 18:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:20

SUBJECT'S LAST NAME: NORTH

DATE/TIME OF TEST: 07/28/2017 19:54:49

Diagnostics Check		OK	19:51	
Air Blank		0.000	19:52	
Control Test		0.077	19:52	
Air Blank		0.000	19:53	
Subject Sample	# 1		19:53	Subject Test Refused
Breath Volume		0.000	19:53	
Air Blank		0.000	19:53	
Control Test		0.076	19:54	
Air Blank		0.000	19:54	
Diagnostics Check		OK	19:54	

Data Download Date/Time: 08/29/2017 14:13

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001115

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVERETT,MATTHEW,A
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: OHARA,K
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 06/07/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: ATKINS

DATE/TIME OF TEST: 07/04/2017 04:16:31

Diagnostics Check		OK	04:10
Air Blank		0.000	04:10
Control Test		0.078	04:11
Air Blank		0.000	04:11
Subject Sample	# 1	0.231	04:12
Breath Volume		1.929	04:12
Air Blank		0.000	04:12
Air Blank		0.000	04:14
Subject Sample	# 2	0.233	04:15
Breath Volume		2.777	04:15
Air Blank		0.000	04:15
Control Test		0.078	04:15
Air Blank		0.000	04:16
Diagnostics Check		OK	04:16

Data Download Date/Time: 07/13/2017 10:41

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001115

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARMER,GLENCHARLES,J
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: FARMER,GLENCHARLES
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 06/07/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: GRADDY

DATE/TIME OF TEST: 07/06/2017 02:10:00

Diagnostics Check		OK	02:07	
Air Blank		0.000	02:07	
Control Test		0.078	02:07	
Air Blank		0.000	02:08	
Subject Sample	# 1		02:08	Subject Test Refused
Breath Volume		0.000	02:08	
Air Blank		0.000	02:09	
Control Test		0.078	02:09	
Air Blank		0.000	02:09	
Diagnostics Check		OK	02:10	

Data Download Date/Time: 07/13/2017 10:41

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001115

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHULZE,J,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: SCHULZE,J
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 06/07/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:52

SUBJECT'S LAST NAME: NAZWORTH

DATE/TIME OF TEST: 07/07/2017 23:19:01

Diagnostics Check		OK	23:12
Air Blank		0.000	23:13
Control Test		0.078	23:13
Air Blank		0.000	23:13
Subject Sample	# 1	0.173	23:14
Breath Volume		2.984	23:14
Air Blank		0.000	23:15
Air Blank		0.000	23:16
Subject Sample	# 2	0.165	23:17
Breath Volume		1.929	23:17
Air Blank		0.000	23:18
Control Test		0.078	23:18
Air Blank		0.000	23:18
Diagnostics Check		OK	23:19

Data Download Date/Time: 07/13/2017 10:41

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001115

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,A,L
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: GONZALEZ,A
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 06/07/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:40

SUBJECT'S LAST NAME: FITZGERALD

DATE/TIME OF TEST: 07/09/2017 21:12:35

Diagnostics Check		OK	21:06
Air Blank		0.000	21:07
Control Test		0.078	21:07
Air Blank		0.000	21:07
Subject Sample	# 1	0.310	21:08
Breath Volume		1.773	21:08
Air Blank		0.000	21:08
Air Blank		0.000	21:10
Subject Sample	# 2	0.299	21:11
Breath Volume		1.953	21:11
Air Blank		0.000	21:11
Control Test		0.078	21:11
Air Blank		0.000	21:12
Diagnostics Check		OK	21:12

Data Download Date/Time: 07/13/2017 10:41

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001115

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KELLER,DAVID,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: KELLER,DAVID
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 06/07/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:48

SUBJECT'S LAST NAME: LYLE

DATE/TIME OF TEST: 07/12/2017 23:14:43

Diagnostics Check		OK	23:08
Air Blank		0.000	23:08
Control Test		0.078	23:09
Air Blank		0.000	23:09
Subject Sample	# 1	0.060	23:10
Breath Volume		2.015	23:10
Air Blank		0.000	23:10
Air Blank		0.000	23:12
Subject Sample	# 2	0.051	23:13
Breath Volume		3.402	23:13
Air Blank		0.000	23:13
Control Test		0.078	23:14
Air Blank		0.000	23:14
Diagnostics Check		OK	23:14

Data Download Date/Time: 07/13/2017 10:41

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001115

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'CONNOR,STEPHEN,L
BREATH TEST OPERATOR AGENCY: Leon County SO
ARRESTING OFFICER: O'CONNOR,STEPHEN
ARRESTING OFFICER AGENCY: Leon County SO
LAST AGENCY INSPECTION: 06/07/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:08

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 07/13/2017 05:39:10

Diagnostics Check		OK	05:32
Air Blank		0.000	05:32
Control Test		0.077	05:33
Air Blank		0.000	05:33
Subject Sample	# 1	0.121	05:34
Breath Volume		2.277	05:34
Air Blank		0.000	05:35
Air Blank		0.000	05:36
Subject Sample	# 2	0.112	05:37
Breath Volume		2.632	05:37
Air Blank		0.000	05:38
Control Test		0.076	05:38
Air Blank		0.000	05:39
Diagnostics Check		OK	05:39

Data Download Date/Time: 07/13/2017 10:41

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001115

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILSON,BRANDON,J
BREATH TEST OPERATOR AGENCY: FL State University
ARRESTING OFFICER: WILSON,BRANDON
ARRESTING OFFICER AGENCY: FL State University
LAST AGENCY INSPECTION: 07/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 07/14/2017 02:39:16

Diagnostics Check		OK	02:30	
Air Blank		0.000	02:31	
Control Test		0.077	02:31	
Air Blank		0.000	02:32	
Subject Sample	# 1		02:35	No Sample Provided
Breath Volume		0.000	02:35	
Air Blank		0.000	02:35	
Air Blank		0.000	02:37	
Subject Sample	# 2		02:37	Subject Test Refused
Breath Volume		0.000	02:37	
Air Blank		0.000	02:38	
Control Test		0.078	02:38	
Air Blank		0.000	02:39	
Diagnostics Check		OK	02:39	

Data Download Date/Time: 08/07/2017 11:56

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001115

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KELLER,DAVID,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: NORTHWAY,G
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 07/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: STROM

DATE/TIME OF TEST: 07/15/2017 03:35:17

Diagnostics Check		OK	03:29
Air Blank		0.000	03:29
Control Test		0.078	03:30
Air Blank		0.000	03:30
Subject Sample	# 1	0.224	03:30
Breath Volume		1.738	03:30
Air Blank		0.000	03:31
Air Blank		0.000	03:33
Subject Sample	# 2	0.213	03:33
Breath Volume		1.750	03:33
Air Blank		0.000	03:34
Control Test		0.078	03:34
Air Blank		0.000	03:35
Diagnostics Check		OK	03:35

Data Download Date/Time: 08/07/2017 11:56

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001115

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BULLOCK,JOSHUA,R
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: BULLOCK,JOSHUA
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 07/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: ROBERTS

DATE/TIME OF TEST: 07/15/2017 04:35:06

Diagnostics Check		OK	04:28
Air Blank		0.000	04:29
Control Test		0.079	04:29
Air Blank		0.000	04:29
Subject Sample	# 1	0.146	04:30
Breath Volume		2.789	04:30
Air Blank		0.000	04:31
Air Blank		0.000	04:32
Subject Sample	# 2	0.137	04:33
Breath Volume		1.605	04:33
Air Blank		0.000	04:34
Control Test		0.079	04:34
Air Blank		0.000	04:35
Diagnostics Check		OK	04:35

Data Download Date/Time: 08/07/2017 11:56

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001115

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVERETT,MATTHEW,A
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: EVERETT,MATTHEW
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 07/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 05:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:31

SUBJECT'S LAST NAME: PETTY

DATE/TIME OF TEST: 07/22/2017 06:01:28

Diagnostics Check		OK	05:54
Air Blank		0.000	05:55
Control Test		0.078	05:55
Air Blank		0.000	05:56
Subject Sample	# 1	0.257	05:56
Breath Volume		2.496	05:56
Air Blank		0.000	05:57
Air Blank		0.000	05:59
Subject Sample	# 2	0.249	05:59
Breath Volume		1.199	05:59
Air Blank		0.000	06:00
Control Test		0.078	06:00
Air Blank		0.000	06:01
Diagnostics Check		OK	06:01

Data Download Date/Time: 08/07/2017 11:56

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001115

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILSON,BRANDON,J
BREATH TEST OPERATOR AGENCY: FL State University
ARRESTING OFFICER: WILSON,BRANDON
ARRESTING OFFICER AGENCY: FL State University
LAST AGENCY INSPECTION: 07/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: MC FARLANE

DATE/TIME OF TEST: 07/23/2017 04:18:21

Diagnostics Check		OK	04:10
Air Blank		0.000	04:10
Control Test		0.078	04:10
Air Blank		0.000	04:11
Subject Sample	# 1	0.014	04:12
Breath Volume		1.683	04:12
Air Blank		0.000	04:13
Air Blank		0.000	04:15
Subject Sample	# 2	0.013	04:16
Breath Volume		2.148	04:16
Air Blank		0.000	04:17
Control Test		0.080	04:17
Air Blank		0.000	04:18
Diagnostics Check		OK	04:18

Data Download Date/Time: 08/07/2017 11:56

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001115

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MATTHEWS,JESSE,L
BREATH TEST OPERATOR AGENCY: Leon County SO
ARRESTING OFFICER: MCKOWEN,KAMERON
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 07/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:26

SUBJECT'S LAST NAME: MCKOWEN

DATE/TIME OF TEST: 07/24/2017 04:02:20

Diagnostics Check		OK	03:55
Air Blank		0.000	03:55
Control Test		0.078	03:55
Air Blank		0.000	03:56
Subject Sample	# 1	0.104	03:57
Breath Volume		2.089	03:57
Air Blank		0.000	03:58
Air Blank		0.000	04:00
Subject Sample	# 2	0.106	04:00
Breath Volume		1.531	04:00
Air Blank		0.000	04:01
Control Test		0.078	04:01
Air Blank		0.000	04:02
Diagnostics Check		OK	04:02

Data Download Date/Time: 08/07/2017 11:56

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001115

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LECKINGER,MICHAEL,E
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: LECKINGER,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 07/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: SCHENDEL

DATE/TIME OF TEST: 07/26/2017 00:58:34

Diagnostics Check		OK	00:52
Air Blank		0.000	00:52
Control Test		0.078	00:53
Air Blank		0.000	00:53
Subject Sample	# 1	0.155	00:54
Breath Volume		1.554	00:54
Air Blank		0.000	00:54
Air Blank		0.000	00:56
Subject Sample	# 2	0.164	00:57
Breath Volume		1.132	00:57
Air Blank		0.000	00:57
Control Test		0.077	00:57
Air Blank		0.000	00:58
Diagnostics Check		OK	00:58

Data Download Date/Time: 08/07/2017 11:56

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001115

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROY,G,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: ROY,G
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 07/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:52

SUBJECT'S LAST NAME: FRANKLIN

DATE/TIME OF TEST: 07/26/2017 02:21:12

Diagnostics Check		OK	02:14
Air Blank		0.000	02:15
Control Test		0.079	02:15
Air Blank		0.000	02:15
Subject Sample	# 1	0.157	02:16
Breath Volume		1.957	02:16
Air Blank		0.000	02:17
Air Blank		0.000	02:19
Subject Sample	# 2	0.157	02:19
Breath Volume		1.433	02:19
Air Blank		0.000	02:20
Control Test		0.079	02:20
Air Blank		0.000	02:21
Diagnostics Check		OK	02:21

Data Download Date/Time: 08/07/2017 11:56

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001115

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRAXTON,D,L
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: BRAXTON,D
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 07/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: MORGAN

DATE/TIME OF TEST: 07/27/2017 00:19:37

Diagnostics Check		OK	00:05	
Air Blank		0.000	00:06	
Control Test		0.078	00:06	
Air Blank		0.000	00:06	
Subject Sample	# 1	0.157	00:09	Volume Not Met
Breath Volume		0.062	00:09	
Air Blank		0.000	00:10	
Air Blank		0.000	00:12	
Subject Sample	# 2	0.159	00:14	
Breath Volume		1.218	00:14	
Air Blank		0.000	00:15	
Air Blank		0.000	00:17	
Subject Sample	# 3	0.156	00:17	
Breath Volume		1.339	00:17	
Air Blank		0.000	00:18	
Control Test		0.076	00:19	
Air Blank		0.000	00:19	
Diagnostics Check		OK	00:19	

Data Download Date/Time: 08/07/2017 11:57

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001115

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KELLER,DAVID,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: KELLER,DAVID
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 07/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:38

SUBJECT'S LAST NAME: DUNCAN

DATE/TIME OF TEST: 07/29/2017 03:09:30

Diagnostics Check		OK	02:59
Air Blank		0.000	02:59
Control Test		0.079	03:00
Air Blank		0.000	03:00
Subject Sample	# 1	0.181	03:01
Breath Volume		1.425	03:01
Air Blank		0.000	03:02
Air Blank		0.000	03:03
Subject Sample	# 2	0.207	03:04
Breath Volume		2.144	03:04
Air Blank		0.000	03:05
Air Blank		0.000	03:07
Subject Sample	# 3	0.204	03:07
Breath Volume		1.566	03:07
Air Blank		0.000	03:08
Control Test		0.078	03:08
Air Blank		0.000	03:09
Diagnostics Check		OK	03:09

Data Download Date/Time: 08/07/2017 11:57

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001115

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRIS,JAMES,J
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: HARRIS,JAMES
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 07/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 07:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:05

SUBJECT'S LAST NAME: STARKEY

DATE/TIME OF TEST: 07/30/2017 08:32:26

Diagnostics Check		OK	08:26
Air Blank		0.000	08:26
Control Test		0.078	08:27
Air Blank		0.000	08:27
Subject Sample	# 1	0.170	08:28
Breath Volume		2.484	08:28
Air Blank		0.000	08:28
Air Blank		0.000	08:30
Subject Sample	# 2	0.169	08:30
Breath Volume		2.234	08:30
Air Blank		0.000	08:31
Control Test		0.078	08:31
Air Blank		0.000	08:32
Diagnostics Check		OK	08:32

Data Download Date/Time: 08/07/2017 11:57

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001118

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARGE,SUSAN,M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BARGE,SUSAN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 06/03/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: MATHIS

DATE/TIME OF TEST: 07/01/2017 02:26:19

Diagnostics Check		OK	02:17	
Air Blank		0.000	02:18	
Control Test		0.079	02:18	
Air Blank		0.000	02:19	
Subject Sample	# 1		02:22	No Sample Provided
Breath Volume		0.000	02:22	
Air Blank		0.000	02:22	
Air Blank		0.000	02:24	
Subject Sample	# 2		02:24	Subject Test Refused
Breath Volume		0.000	02:24	
Air Blank		0.000	02:25	
Control Test		0.081	02:25	
Air Blank		0.000	02:26	
Diagnostics Check		OK	02:26	

Data Download Date/Time: 07/25/2017 14:28

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001118

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARGE,SUSAN,M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BARGE,SUSAN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 06/03/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 17:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:17

SUBJECT'S LAST NAME: TRAWICK

DATE/TIME OF TEST: 07/01/2017 18:29:36

Diagnostics Check		OK	18:22
Air Blank		0.000	18:23
Control Test		0.080	18:23
Air Blank		0.000	18:23
Subject Sample	# 1	0.086	18:24
Breath Volume		3.429	18:24
Air Blank		0.000	18:25
Air Blank		0.000	18:27
Subject Sample	# 2	0.088	18:28
Breath Volume		3.847	18:28
Air Blank		0.000	18:28
Control Test		0.080	18:28
Air Blank		0.000	18:29
Diagnostics Check		OK	18:29

Data Download Date/Time: 07/25/2017 14:28

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001118

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARGE,SUSAN,M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: WARD,RON
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 06/03/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: WALKER

DATE/TIME OF TEST: 07/09/2017 01:03:19

Diagnostics Check		OK	00:55
Air Blank		0.000	00:56
Control Test		0.080	00:56
Air Blank		0.000	00:57
Subject Sample	# 1	0.025	00:58
Breath Volume		2.136	00:58
Air Blank		0.000	00:59
Air Blank		0.000	01:00
Subject Sample	# 2	0.024	01:01
Breath Volume		2.179	01:01
Air Blank		0.000	01:02
Control Test		0.080	01:02
Air Blank		0.000	01:03
Diagnostics Check		OK	01:03

Data Download Date/Time: 07/25/2017 14:28

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001118

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GENTRY,JONATHAN,P
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: GENTRY,JONATHAN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 06/03/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 11:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:16

SUBJECT'S LAST NAME: MILLS

DATE/TIME OF TEST: 07/24/2017 00:43:06

Diagnostics Check		OK	00:36
Air Blank		0.000	00:36
Control Test		0.080	00:37
Air Blank		0.000	00:37
Subject Sample	# 1	0.202	00:38
Breath Volume		2.796	00:38
Air Blank		0.000	00:39
Air Blank		0.000	00:40
Subject Sample	# 2	0.196	00:41
Breath Volume		3.347	00:41
Air Blank		0.000	00:42
Control Test		0.080	00:42
Air Blank		0.000	00:43
Diagnostics Check		OK	00:43

Data Download Date/Time: 07/25/2017 14:28

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001119

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,DAILYN,
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: ZIGLER,R
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 07/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 437738
DRY GAS EXPIRY DATE: 06/08/2019
VIOLATION CODE: DUI
ARREST TIME: 09:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:10

SUBJECT'S LAST NAME: MONTROSE

DATE/TIME OF TEST: 07/15/2017 10:36:36

Diagnostics Check		OK	10:30
Air Blank		0.000	10:30
Control Test		0.081	10:31
Air Blank		0.000	10:31
Subject Sample	# 1	0.253	10:32
Breath Volume		1.144	10:32
Air Blank		0.000	10:32
Air Blank		0.000	10:34
Subject Sample	# 2	0.257	10:35
Breath Volume		2.394	10:35
Air Blank		0.000	10:35
Control Test		0.079	10:36
Air Blank		0.000	10:36
Diagnostics Check		OK	10:36

Data Download Date/Time: 08/29/2017 17:52

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001119

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HILDEBRANDT,PATRICK,A
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: NUNEZ,ANDREW
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 07/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 437738
DRY GAS EXPIRY DATE: 06/08/2019
VIOLATION CODE: DUI
ARREST TIME: 23:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:01

SUBJECT'S LAST NAME: LONG

DATE/TIME OF TEST: 07/29/2017 00:36:00

Diagnostics Check		OK	00:29
Air Blank		0.000	00:29
Control Test		0.080	00:29
Air Blank		0.000	00:30
Subject Sample	# 1	0.145	00:31
Breath Volume		1.984	00:31
Air Blank		0.000	00:31
Air Blank		0.000	00:33
Subject Sample	# 2	0.133	00:34
Breath Volume		1.785	00:34
Air Blank		0.000	00:35
Control Test		0.080	00:35
Air Blank		0.000	00:35
Diagnostics Check		OK	00:36

Data Download Date/Time: 08/29/2017 17:52

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001120

INSTRUMENT REGISTERED TO: FHP TROOP K
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAUGHIE,DARIN,A
BREATH TEST OPERATOR AGENCY: FHP Troop K
ARRESTING OFFICER: VARON,MARCOS
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 366849
DRY GAS EXPIRY DATE: 03/30/2019
VIOLATION CODE: DUI
ARREST TIME: 05:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:15

SUBJECT'S LAST NAME: MONTALVO

DATE/TIME OF TEST: 07/01/2017 06:03:49

Diagnostics Check		OK	05:58	
Air Blank		0.000	05:59	
Control Test		0.080	05:59	
Air Blank		0.000	06:00	
Subject Sample	# 1	0.176	06:02	
Breath Volume		1.898	06:02	
Air Blank			06:03	Purge Fail
Air Blank		0.000	06:03	

Data Download Date/Time: 07/07/2017 01:51

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001120

INSTRUMENT REGISTERED TO: FHP TROOP K
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAUGHIE,DARIN,A
BREATH TEST OPERATOR AGENCY: FHP Troop K
ARRESTING OFFICER: VARON,MARCOS
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 366849
DRY GAS EXPIRY DATE: 03/30/2019
VIOLATION CODE: DUI
ARREST TIME: 05:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:15

SUBJECT'S LAST NAME: MONTALVO

DATE/TIME OF TEST: 07/01/2017 06:13:38

Diagnostics Check		OK	06:06
Air Blank		0.000	06:06
Control Test		0.079	06:07
Air Blank		0.000	06:07
Subject Sample	# 1	0.170	06:08
Breath Volume		1.640	06:08
Air Blank		0.000	06:09
Air Blank		0.000	06:11
Subject Sample	# 2	0.168	06:12
Breath Volume		1.496	06:12
Air Blank		0.000	06:12
Control Test		0.079	06:13
Air Blank		0.000	06:13
Diagnostics Check		OK	06:13

Data Download Date/Time: 07/07/2017 01:51

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001120

INSTRUMENT REGISTERED TO: FHP TROOP K
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASTILLO,JOSUET,
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: CASTILLO,JOSUET
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 366849
DRY GAS EXPIRY DATE: 03/30/2019
VIOLATION CODE: DUI
ARREST TIME: 03:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 07/04/2017 04:06:23

Diagnostics Check		OK	04:01	
Air Blank		0.000	04:01	
Control Test		0.080	04:01	
Air Blank		0.000	04:02	
Subject Sample	# 1		04:05	Subject Test Refused
Breath Volume		0.000	04:05	
Air Blank		0.000	04:05	
Control Test		0.080	04:05	
Air Blank		0.000	04:06	
Diagnostics Check		OK	04:06	

Data Download Date/Time: 07/07/2017 01:51

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001130

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RITTENOUR,COREY,T
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: RITTENOUR,COREY
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 06/20/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 01:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: MORGESE

DATE/TIME OF TEST: 07/01/2017 02:31:44

Diagnostics Check		OK	02:23
Air Blank		0.000	02:23
Control Test		0.078	02:23
Air Blank		0.000	02:24
Subject Sample	# 1	0.180	02:26
Breath Volume		1.597	02:26
Air Blank		0.000	02:26
Air Blank		0.000	02:28
Subject Sample	# 2	0.181	02:30
Breath Volume		2.355	02:30
Air Blank		0.000	02:30
Control Test		0.078	02:31
Air Blank		0.000	02:31
Diagnostics Check		OK	02:31

Data Download Date/Time: 07/26/2017 08:30

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001130

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PECK,KEITH,D
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: PECK,KEITH
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 06/20/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 22:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:48

SUBJECT'S LAST NAME: TATE

DATE/TIME OF TEST: 07/13/2017 23:11:50

Diagnostics Check		OK	23:08	
Air Blank		0.000	23:09	
Control Test		0.078	23:09	
Air Blank		0.000	23:10	
Subject Sample	# 1		23:10	Subject Test Refused
Breath Volume		0.000	23:10	
Air Blank		0.000	23:10	
Control Test		0.079	23:11	
Air Blank		0.000	23:11	
Diagnostics Check		OK	23:11	

Data Download Date/Time: 07/26/2017 08:30

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001130

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PECK,KEITH,D
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: PECK,KEITH
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 06/20/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:34

SUBJECT'S LAST NAME: BUCKLEY

DATE/TIME OF TEST: 07/19/2017 03:02:38

Diagnostics Check		OK	02:54
Air Blank		0.000	02:55
Control Test		0.078	02:55
Air Blank		0.000	02:56
Subject Sample	# 1	0.000	02:56
Breath Volume		2.121	02:56
Air Blank		0.000	02:57
Air Blank		0.000	02:59
Subject Sample	# 2	0.000	03:01
Breath Volume		2.429	03:01
Air Blank		0.000	03:01
Control Test		0.078	03:02
Air Blank		0.000	03:02
Diagnostics Check		OK	03:02

Data Download Date/Time: 07/26/2017 08:30

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001130

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PECK,KEITH,D
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: PECK,KEITH
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 06/20/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 00:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:42

SUBJECT'S LAST NAME: HARDY

DATE/TIME OF TEST: 07/24/2017 01:14:47

Diagnostics Check		OK	01:11	
Air Blank		0.000	01:12	
Control Test		0.078	01:12	
Air Blank		0.000	01:13	
Subject Sample	# 1		01:13	Subject Test Refused
Breath Volume		0.000	01:13	
Air Blank		0.000	01:13	
Control Test		0.078	01:14	
Air Blank		0.000	01:14	
Diagnostics Check		OK	01:14	

Data Download Date/Time: 07/26/2017 08:30

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001130

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,MATTHEW,A
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: DEMARCO,JAMES
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 07/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 00:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:56

SUBJECT'S LAST NAME: MITTO

DATE/TIME OF TEST: 07/31/2017 01:37:44

Diagnostics Check		OK	01:31
Air Blank		0.000	01:31
Control Test		0.079	01:32
Air Blank		0.000	01:32
Subject Sample	# 1	0.014	01:33
Breath Volume		1.707	01:33
Air Blank		0.000	01:33
Air Blank		0.000	01:35
Subject Sample	# 2	0.015	01:36
Breath Volume		1.824	01:36
Air Blank		0.000	01:36
Control Test		0.079	01:37
Air Blank		0.000	01:37
Diagnostics Check		OK	01:37

Data Download Date/Time: 08/29/2017 14:00

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001131

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOOLEYHAN,JARETT,G
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: CLOUTIER,KYLE
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 06/20/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 02:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:51

SUBJECT'S LAST NAME: GOODMAN

DATE/TIME OF TEST: 07/19/2017 03:22:56

Diagnostics Check		OK	03:16
Air Blank		0.000	03:17
Control Test		0.079	03:17
Air Blank		0.000	03:18
Subject Sample	# 1	0.162	03:18
Breath Volume		2.425	03:18
Air Blank		0.000	03:19
Air Blank		0.000	03:21
Subject Sample	# 2	0.164	03:21
Breath Volume		2.558	03:21
Air Blank		0.000	03:22
Control Test		0.080	03:22
Air Blank		0.000	03:22
Diagnostics Check		OK	03:22

Data Download Date/Time: 07/26/2017 09:36

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.