

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001275

INSTRUMENT REGISTERED TO: HIALEAH GARDENS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZORRILLA,J,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: ZORRILLA,J
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 08/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 11:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:25

SUBJECT'S LAST NAME: RESTREPOCHAVERRA

DATE/TIME OF TEST: 09/13/2017 13:19:19

Diagnostics Check		OK	13:05	
Air Blank		0.000	13:05	
Control Test		0.079	13:06	
Air Blank		0.000	13:06	
Subject Sample	# 1	0.000	13:09	Volume Not Met
Breath Volume		0.050	13:09	
Air Blank		0.000	13:10	
Air Blank		0.000	13:12	
Subject Sample	# 2	0.000	13:15	
Breath Volume		1.281	13:15	
Air Blank		0.000	13:15	
Air Blank		0.000	13:17	
Subject Sample	# 3	0.000	13:17	
Breath Volume		1.914	13:17	
Air Blank		0.000	13:18	
Control Test		0.077	13:18	
Air Blank		0.000	13:19	
Diagnostics Check		OK	13:19	

Data Download Date/Time: 10/02/2017 13:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: CLOSIUS,RICHARD
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 08/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:54

SUBJECT'S LAST NAME: TORRA

DATE/TIME OF TEST: 09/01/2017 02:22:47

Diagnostics Check	OK	02:19	
Air Blank	0.000	02:20	
Control Test	0.077	02:20	
Air Blank	0.000	02:21	
Subject Sample	# 1	02:21	Subject Test Refused
Breath Volume	0.000	02:21	
Air Blank	0.000	02:21	
Control Test	0.077	02:22	
Air Blank	0.000	02:22	
Diagnostics Check	OK	02:22	

Data Download Date/Time: 09/04/2017 10:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: VARGAS,DAEL
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 09/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 09/18/2017 01:39:08

Diagnostics Check	OK	01:32
Air Blank	0.000	01:33
Control Test	0.079	01:33
Air Blank	0.000	01:34
Subject Sample # 1	0.170	01:34
Breath Volume	2.199	01:34
Air Blank	0.000	01:35
Air Blank	0.000	01:37
Subject Sample # 2	0.182	01:37
Breath Volume	1.507	01:37
Air Blank	0.000	01:38
Control Test	0.079	01:38
Air Blank	0.000	01:39
Diagnostics Check	OK	01:39

Data Download Date/Time: 10/13/2017 10:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: CLOSIUS,RICHARD
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 09/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:16

SUBJECT'S LAST NAME: GONZALEZ ABREU

DATE/TIME OF TEST: 09/21/2017 02:52:16

Diagnostics Check	OK	02:42
Air Blank	0.000	02:43
Control Test	0.079	02:43
Air Blank	0.000	02:43
Subject Sample # 1	0.179	02:46
Breath Volume	1.121	02:46
Air Blank	0.000	02:47
Air Blank	0.000	02:49
Subject Sample # 2	0.171	02:50
Breath Volume	1.265	02:50
Air Blank	0.000	02:51
Control Test	0.076	02:51
Air Blank	0.000	02:52
Diagnostics Check	OK	02:52

Data Download Date/Time: 10/13/2017 10:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 09/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:03

SUBJECT'S LAST NAME: COLL

DATE/TIME OF TEST: 09/21/2017 05:29:15

Diagnostics Check	OK	05:23
Air Blank	0.000	05:23
Control Test	0.078	05:23
Air Blank	0.000	05:24
Subject Sample # 1	0.083	05:24
Breath Volume	2.023	05:24
Air Blank	0.000	05:25
Air Blank	0.000	05:27
Subject Sample # 2	0.079	05:27
Breath Volume	1.429	05:27
Air Blank	0.000	05:28
Control Test	0.079	05:28
Air Blank	0.000	05:29
Diagnostics Check	OK	05:29

Data Download Date/Time: 10/13/2017 10:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: CLOSIUS,RICHARD
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 09/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:38

SUBJECT'S LAST NAME: DUBE

DATE/TIME OF TEST: 09/29/2017 04:05:34

Diagnostics Check		OK	03:58
Air Blank		0.000	03:59
Control Test		0.079	03:59
Air Blank		0.000	04:00
Subject Sample	# 1	0.106	04:00
Breath Volume		1.332	04:00
Air Blank		0.000	04:01
Air Blank		0.000	04:03
Subject Sample	# 2	0.108	04:03
Breath Volume		1.859	04:03
Air Blank		0.000	04:04
Control Test		0.078	04:04
Air Blank		0.000	04:05
Diagnostics Check		OK	04:05

Data Download Date/Time: 10/13/2017 10:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AUGUSTINE,GRANT,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CAMACHO,K
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 08/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:52

SUBJECT'S LAST NAME: VILCHIS LOPEZ

DATE/TIME OF TEST: 09/01/2017 00:24:09

Diagnostics Check		OK	00:12	
Air Blank		0.000	00:13	
Control Test		0.080	00:13	
Air Blank		0.000	00:13	
Subject Sample	# 1	0.165	00:17	Volume Not Met
Breath Volume		0.269	00:17	
Air Blank		0.000	00:17	
Air Blank		0.000	00:19	
Subject Sample	# 2		00:22	Subject Test Refused
Breath Volume		0.382	00:22	
Air Blank		0.000	00:23	
Control Test		0.078	00:23	
Air Blank		0.000	00:24	
Diagnostics Check		OK	00:24	

Data Download Date/Time: 09/12/2017 12:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AUGUSTINE, GRANT,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHITE, J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:27

SUBJECT'S LAST NAME: STOKES

DATE/TIME OF TEST: 09/01/2017 02:54:27

Diagnostics Check	OK	02:48
Air Blank	0.000	02:48
Control Test	0.079	02:48
Air Blank	0.000	02:49
Subject Sample # 1	0.150	02:50
Breath Volume	2.582	02:50
Air Blank	0.000	02:50
Air Blank	0.000	02:52
Subject Sample # 2	0.148	02:52
Breath Volume	2.792	02:52
Air Blank	0.000	02:53
Control Test	0.079	02:53
Air Blank	0.000	02:54
Diagnostics Check	OK	02:54

Data Download Date/Time: 09/12/2017 12:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AUGUSTINE, GRANT,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MOELLER, R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: KATTREH

DATE/TIME OF TEST: 09/01/2017 23:17:06

Diagnostics Check	OK	23:10
Air Blank	0.000	23:11
Control Test	0.079	23:11
Air Blank	0.000	23:11
Subject Sample # 1	0.126	23:12
Breath Volume	1.808	23:12
Air Blank	0.000	23:13
Air Blank	0.000	23:15
Subject Sample # 2	0.129	23:15
Breath Volume	1.679	23:15
Air Blank	0.000	23:16
Control Test	0.079	23:16
Air Blank	0.000	23:17
Diagnostics Check	OK	23:17

Data Download Date/Time: 09/12/2017 12:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AUGUSTINE, GRANT,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FISETTE, J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:52

SUBJECT'S LAST NAME: MINIX

DATE/TIME OF TEST: 09/02/2017 01:19:15

Diagnostics Check	OK	01:13
Air Blank	0.000	01:13
Control Test	0.079	01:13
Air Blank	0.000	01:14
Subject Sample # 1	0.063	01:14
Breath Volume	1.652	01:14
Air Blank	0.000	01:15
Air Blank	0.000	01:17
Subject Sample # 2	0.059	01:17
Breath Volume	1.847	01:17
Air Blank	0.000	01:18
Control Test	0.079	01:18
Air Blank	0.000	01:19
Diagnostics Check	OK	01:19

Data Download Date/Time: 09/12/2017 12:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AUGUSTINE, GRANT,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MOELLER, R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:24

SUBJECT'S LAST NAME: YOUNG

DATE/TIME OF TEST: 09/02/2017 03:50:29

Diagnostics Check	OK	03:44
Air Blank	0.000	03:45
Control Test	0.079	03:45
Air Blank	0.000	03:45
Subject Sample # 1	0.000	03:46
Breath Volume	2.472	03:46
Air Blank	0.000	03:46
Air Blank	0.000	03:48
Subject Sample # 2	0.000	03:49
Breath Volume	2.285	03:49
Air Blank	0.000	03:49
Control Test	0.080	03:49
Air Blank	0.000	03:50
Diagnostics Check	OK	03:50

Data Download Date/Time: 09/12/2017 12:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BROWN,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:52

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 09/02/2017 22:15:49

Diagnostics Check	OK	22:12	
Air Blank	0.000	22:13	
Control Test	0.079	22:13	
Air Blank	0.000	22:14	
Subject Sample	# 1	22:14	Subject Test Refused
Breath Volume	0.000	22:14	
Air Blank	0.000	22:14	
Control Test	0.079	22:15	
Air Blank	0.000	22:15	
Diagnostics Check	OK	22:15	

Data Download Date/Time: 09/12/2017 12:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BRYANT,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:43

SUBJECT'S LAST NAME: WIMBERLY

DATE/TIME OF TEST: 09/03/2017 02:07:52

Diagnostics Check	OK	02:03	
Air Blank	0.000	02:04	
Control Test	0.080	02:04	
Air Blank	0.000	02:05	
Subject Sample	# 1	02:06	Subject Test Refused
Breath Volume	0.000	02:06	
Air Blank	0.000	02:06	
Control Test	0.079	02:07	
Air Blank	0.000	02:07	
Diagnostics Check	OK	02:07	

Data Download Date/Time: 09/12/2017 12:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AUGUSTINE, GRANT,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HEALY, G
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 08/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:13

SUBJECT'S LAST NAME: HEDGEMOND

DATE/TIME OF TEST: 09/03/2017 03:41:11

Diagnostics Check	OK	03:33
Air Blank	0.000	03:34
Control Test	0.079	03:34
Air Blank	0.000	03:35
Subject Sample # 1	0.212	03:36
Breath Volume	3.664	03:36
Air Blank	0.000	03:37
Air Blank	0.000	03:39
Subject Sample # 2	0.209	03:39
Breath Volume	3.882	03:39
Air Blank	0.000	03:40
Control Test	0.079	03:40
Air Blank	0.000	03:41
Diagnostics Check	OK	03:41

Data Download Date/Time: 09/12/2017 12:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AUGUSTINE,GRANT,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: KAMPPI,D
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 08/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:44

SUBJECT'S LAST NAME: IVAN

DATE/TIME OF TEST: 09/03/2017 05:08:07

Diagnostics Check	OK	05:05	
Air Blank	0.000	05:05	
Control Test	0.079	05:06	
Air Blank	0.000	05:06	
Subject Sample	# 1	05:06	Subject Test Refused
Breath Volume	0.000	05:06	
Air Blank	0.000	05:07	
Control Test	0.080	05:07	
Air Blank	0.000	05:08	
Diagnostics Check	OK	05:08	

Data Download Date/Time: 09/12/2017 12:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AUGUSTINE, GRANT,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JOINER, D
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 08/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:19

SUBJECT'S LAST NAME: STOKES

DATE/TIME OF TEST: 09/03/2017 05:46:01

Diagnostics Check	OK	05:39
Air Blank	0.000	05:40
Control Test	0.080	05:40
Air Blank	0.000	05:40
Subject Sample # 1	0.126	05:41
Breath Volume	1.535	05:41
Air Blank	0.000	05:42
Air Blank	0.000	05:44
Subject Sample # 2	0.122	05:44
Breath Volume	1.277	05:44
Air Blank	0.000	05:45
Control Test	0.079	05:45
Air Blank	0.000	05:45
Diagnostics Check	OK	05:46

Data Download Date/Time: 09/12/2017 12:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HEALY,G
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 08/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:53

SUBJECT'S LAST NAME: BOLTE

DATE/TIME OF TEST: 09/04/2017 00:20:12

Diagnostics Check		OK	00:13
Air Blank		0.000	00:14
Control Test		0.079	00:14
Air Blank		0.000	00:15
Subject Sample	# 1	0.230	00:15
Breath Volume		2.468	00:15
Air Blank		0.000	00:16
Air Blank		0.000	00:18
Subject Sample	# 2	0.225	00:18
Breath Volume		2.230	00:18
Air Blank		0.000	00:19
Control Test		0.079	00:19
Air Blank		0.000	00:20
Diagnostics Check		OK	00:20

Data Download Date/Time: 09/12/2017 12:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BYRNE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:49

SUBJECT'S LAST NAME: JEREZ SANCHEZ

DATE/TIME OF TEST: 09/04/2017 01:16:19

Diagnostics Check	OK	01:09
Air Blank	0.000	01:10
Control Test	0.079	01:10
Air Blank	0.000	01:11
Subject Sample # 1	0.196	01:12
Breath Volume	1.820	01:12
Air Blank	0.000	01:12
Air Blank	0.000	01:14
Subject Sample # 2	0.205	01:14
Breath Volume	1.734	01:14
Air Blank	0.000	01:15
Control Test	0.080	01:15
Air Blank	0.000	01:16
Diagnostics Check	OK	01:16

Data Download Date/Time: 09/12/2017 12:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HEALY,G
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 08/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:36

SUBJECT'S LAST NAME: AMBROSE

DATE/TIME OF TEST: 09/04/2017 02:03:40

Diagnostics Check		OK	01:57
Air Blank		0.000	01:57
Control Test		0.079	01:58
Air Blank		0.000	01:58
Subject Sample	# 1	0.259	01:59
Breath Volume		2.261	01:59
Air Blank		0.000	01:59
Air Blank		0.000	02:01
Subject Sample	# 2	0.247	02:02
Breath Volume		2.593	02:02
Air Blank		0.000	02:02
Control Test		0.079	02:03
Air Blank		0.000	02:03
Diagnostics Check		OK	02:03

Data Download Date/Time: 09/12/2017 12:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DRAGOMIR RAMINEZ,I
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 08/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:26

SUBJECT'S LAST NAME: CHAMBERS

DATE/TIME OF TEST: 09/05/2017 05:52:39

Diagnostics Check	OK	05:46
Air Blank	0.000	05:47
Control Test	0.079	05:47
Air Blank	0.000	05:47
Subject Sample # 1	0.236	05:48
Breath Volume	2.414	05:48
Air Blank	0.000	05:48
Air Blank	0.000	05:50
Subject Sample # 2	0.232	05:51
Breath Volume	1.800	05:51
Air Blank	0.000	05:51
Control Test	0.079	05:52
Air Blank	0.000	05:52
Diagnostics Check	OK	05:52

Data Download Date/Time: 09/12/2017 12:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BRYANT,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:51

SUBJECT'S LAST NAME: MORMAN

DATE/TIME OF TEST: 09/05/2017 23:15:24

Diagnostics Check	OK	23:12	
Air Blank	0.000	23:13	
Control Test	0.081	23:13	
Air Blank	0.000	23:13	
Subject Sample	# 1	23:14	Subject Test Refused
Breath Volume	0.000	23:14	
Air Blank	0.000	23:14	
Control Test	0.079	23:14	
Air Blank	0.000	23:15	
Diagnostics Check	OK	23:15	

Data Download Date/Time: 09/12/2017 12:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LANGSTON,B
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:16

SUBJECT'S LAST NAME: EUNICE

DATE/TIME OF TEST: 09/06/2017 02:43:34

Diagnostics Check	OK	02:37
Air Blank	0.000	02:38
Control Test	0.080	02:38
Air Blank	0.000	02:38
Subject Sample # 1	0.000	02:39
Breath Volume	2.730	02:39
Air Blank	0.000	02:39
Air Blank	0.000	02:41
Subject Sample # 2	0.000	02:42
Breath Volume	3.074	02:42
Air Blank	0.000	02:42
Control Test	0.080	02:42
Air Blank	0.000	02:43
Diagnostics Check	OK	02:43

Data Download Date/Time: 09/12/2017 12:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENDRIX,D,K
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: KINARD,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 08:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:52

SUBJECT'S LAST NAME: HOGAN

DATE/TIME OF TEST: 09/06/2017 10:21:04

Diagnostics Check	OK	10:12
Air Blank	0.000	10:12
Control Test	0.080	10:12
Air Blank	0.000	10:13
Subject Sample # 1	0.321	10:15
Breath Volume	1.929	10:15
Air Blank	0.000	10:16
Air Blank	0.000	10:17
Subject Sample # 2	0.324	10:19
Breath Volume	1.851	10:19
Air Blank	0.000	10:20
Control Test	0.079	10:20
Air Blank	0.000	10:20
Diagnostics Check	OK	10:21

Data Download Date/Time: 09/12/2017 12:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AUGUSTINE, GRANT,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HILLER, J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: NUNG

DATE/TIME OF TEST: 09/09/2017 22:46:32

Diagnostics Check	OK	22:40
Air Blank	0.000	22:41
Control Test	0.080	22:41
Air Blank	0.000	22:41
Subject Sample # 1	0.149	22:42
Breath Volume	1.906	22:42
Air Blank	0.000	22:42
Air Blank	0.000	22:44
Subject Sample # 2	0.140	22:45
Breath Volume	2.531	22:45
Air Blank	0.000	22:45
Control Test	0.079	22:45
Air Blank	0.000	22:46
Diagnostics Check	OK	22:46

Data Download Date/Time: 09/12/2017 12:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ZURN,E
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: OTHER
ARREST TIME: 20:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:37

SUBJECT'S LAST NAME: HELLER

DATE/TIME OF TEST: 09/11/2017 23:03:12

Diagnostics Check	OK	22:57
Air Blank	0.000	22:57
Control Test	0.080	22:58
Air Blank	0.000	22:58
Subject Sample # 1	0.079	22:58
Breath Volume	2.097	22:58
Air Blank	0.000	22:59
Air Blank	0.000	23:01
Subject Sample # 2	0.080	23:01
Breath Volume	2.261	23:01
Air Blank	0.000	23:02
Control Test	0.080	23:02
Air Blank	0.000	23:03
Diagnostics Check	OK	23:03

Data Download Date/Time: 09/12/2017 12:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MURRIELL,E,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BARRY,H
ARRESTING OFFICER AGENCY: FFWCC (Lake City)
LAST AGENCY INSPECTION: 08/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: BUI
ARREST TIME: 18:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:23

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 09/03/2017 19:47:08

Diagnostics Check	OK	19:43	
Air Blank	0.000	19:44	
Control Test	0.079	19:44	
Air Blank	0.000	19:45	
Subject Sample	# 1	19:45	Subject Test Refused
Breath Volume	0.000	19:45	
Air Blank	0.000	19:46	
Control Test	0.080	19:46	
Air Blank	0.000	19:47	
Diagnostics Check	OK	19:47	

Data Download Date/Time: 09/12/2017 12:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MURRIELL,E,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BRYANT,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:08

SUBJECT'S LAST NAME: MITCHELL

DATE/TIME OF TEST: 09/03/2017 22:35:37

Diagnostics Check	OK	22:28
Air Blank	0.000	22:29
Control Test	0.080	22:29
Air Blank	0.000	22:30
Subject Sample # 1	0.038	22:31
Breath Volume	1.687	22:31
Air Blank	0.000	22:31
Air Blank	0.000	22:33
Subject Sample # 2	0.035	22:34
Breath Volume	1.804	22:34
Air Blank	0.000	22:34
Control Test	0.079	22:35
Air Blank	0.000	22:35
Diagnostics Check	OK	22:35

Data Download Date/Time: 09/12/2017 12:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,J,T
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BRYANT,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:02

SUBJECT'S LAST NAME: ONEAL

DATE/TIME OF TEST: 09/04/2017 22:28:46

Diagnostics Check	OK	22:22
Air Blank	0.000	22:23
Control Test	0.078	22:23
Air Blank	0.000	22:24
Subject Sample # 1	0.277	22:24
Breath Volume	1.507	22:24
Air Blank	0.000	22:25
Air Blank	0.000	22:26
Subject Sample # 2	0.273	22:27
Breath Volume	1.625	22:27
Air Blank	0.000	22:27
Control Test	0.080	22:28
Air Blank	0.000	22:28
Diagnostics Check	OK	22:28

Data Download Date/Time: 09/12/2017 12:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAYLOR,TIMOTHY,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: GARCIA,J.R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 17:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:38

SUBJECT'S LAST NAME: HENRY

DATE/TIME OF TEST: 09/09/2017 19:01:33

Diagnostics Check	OK	18:58	
Air Blank	0.000	18:59	
Control Test	0.081	18:59	
Air Blank	0.000	18:59	
Subject Sample	# 1	19:00	Subject Test Refused
Breath Volume	0.000	19:00	
Air Blank	0.000	19:00	
Control Test	0.080	19:00	
Air Blank	0.000	19:01	
Diagnostics Check	OK	19:01	

Data Download Date/Time: 09/12/2017 12:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAYLOR,TIMOTHY,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: STANLEY,T.L
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 17:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:26

SUBJECT'S LAST NAME: TASSEFF

DATE/TIME OF TEST: 09/09/2017 19:53:33

Diagnostics Check	OK	19:47
Air Blank	0.000	19:47
Control Test	0.080	19:47
Air Blank	0.000	19:48
Subject Sample # 1	0.257	19:49
Breath Volume	3.894	19:49
Air Blank	0.000	19:49
Air Blank	0.000	19:51
Subject Sample # 2	0.255	19:52
Breath Volume	3.648	19:52
Air Blank	0.000	19:52
Control Test	0.078	19:52
Air Blank	0.000	19:53
Diagnostics Check	OK	19:53

Data Download Date/Time: 09/12/2017 12:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAYLOR,TIMOTHY,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MOELLER,R.H
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:11

SUBJECT'S LAST NAME: KOPSO

DATE/TIME OF TEST: 09/09/2017 21:34:53

Diagnostics Check	OK	21:32	
Air Blank	0.000	21:32	
Control Test	0.079	21:32	
Air Blank	0.000	21:33	
Subject Sample	# 1	21:33	Subject Test Refused
Breath Volume	0.000	21:33	
Air Blank	0.000	21:33	
Control Test	0.079	21:34	
Air Blank	0.000	21:34	
Diagnostics Check	OK	21:34	

Data Download Date/Time: 09/12/2017 12:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,J,T
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FISETTE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 17:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:33

SUBJECT'S LAST NAME: MARKOWSKI

DATE/TIME OF TEST: 09/10/2017 19:09:56

Diagnostics Check	OK	19:07	
Air Blank	0.000	19:07	
Control Test	0.079	19:07	
Air Blank	0.000	19:08	
Subject Sample	# 1	19:08	Subject Test Refused
Breath Volume	0.000	19:08	
Air Blank	0.000	19:09	
Control Test	0.080	19:09	
Air Blank	0.000	19:09	
Diagnostics Check	OK	19:09	

Data Download Date/Time: 09/12/2017 12:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,J,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: GARCIA,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: CRIVERO

DATE/TIME OF TEST: 09/10/2017 22:59:50

Diagnostics Check	OK	22:53
Air Blank	0.000	22:53
Control Test	0.078	22:53
Air Blank	0.000	22:54
Subject Sample # 1	0.187	22:55
Breath Volume	3.093	22:55
Air Blank	0.000	22:55
Air Blank	0.000	22:57
Subject Sample # 2	0.182	22:58
Breath Volume	2.621	22:58
Air Blank	0.000	22:58
Control Test	0.079	22:59
Air Blank	0.000	22:59
Diagnostics Check	OK	22:59

Data Download Date/Time: 09/12/2017 12:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MOELLER,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:34

SUBJECT'S LAST NAME: ALVAREZ

DATE/TIME OF TEST: 09/12/2017 04:00:36

Diagnostics Check	OK	03:54
Air Blank	0.000	03:54
Control Test	0.080	03:55
Air Blank	0.000	03:55
Subject Sample # 1	0.102	03:56
Breath Volume	1.398	03:56
Air Blank	0.000	03:56
Air Blank	0.000	03:58
Subject Sample # 2	0.108	03:59
Breath Volume	1.222	03:59
Air Blank	0.000	03:59
Control Test	0.079	04:00
Air Blank	0.000	04:00
Diagnostics Check	OK	04:00

Data Download Date/Time: 09/12/2017 12:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: TORRES,W
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 09/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: MACKKEY

DATE/TIME OF TEST: 09/13/2017 03:46:37

Diagnostics Check		OK	03:40
Air Blank		0.000	03:40
Control Test		0.079	03:41
Air Blank		0.000	03:41
Subject Sample	# 1	0.134	03:42
Breath Volume		1.816	03:42
Air Blank		0.000	03:42
Air Blank		0.000	03:44
Subject Sample	# 2	0.137	03:45
Breath Volume		1.691	03:45
Air Blank		0.000	03:45
Control Test		0.080	03:46
Air Blank		0.000	03:46
Diagnostics Check		OK	03:46

Data Download Date/Time: 09/28/2017 11:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAYLOR,TIMOTHY,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LAZINSKY,N.P
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 09/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 12:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:24

SUBJECT'S LAST NAME: LINDEMAN

DATE/TIME OF TEST: 09/14/2017 17:01:27

Diagnostics Check	OK	16:54	
Air Blank	0.000	16:55	
Control Test	0.080	16:55	
Air Blank	0.000	16:56	
Subject Sample # 1	0.238	16:56	
Breath Volume	2.085	16:56	
Air Blank	0.000	16:57	
Air Blank	0.000	16:59	
Subject Sample # 2		16:59	Slope Not Met
Breath Volume	1.156	16:59	
Air Blank	0.000	17:00	
Control Test	0.079	17:00	
Air Blank	0.000	17:01	
Diagnostics Check	OK	17:01	

Data Download Date/Time: 09/28/2017 11:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAYLOR,TIMOTHY,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LAPINSKY,N.P
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 09/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 12:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:01

SUBJECT'S LAST NAME: LINDEMAN

DATE/TIME OF TEST: 09/14/2017 17:28:15

Diagnostics Check	OK	17:21
Air Blank	0.000	17:22
Control Test	0.079	17:22
Air Blank	0.000	17:23
Subject Sample # 1	0.229	17:23
Breath Volume	2.125	17:23
Air Blank	0.000	17:24
Air Blank	0.000	17:26
Subject Sample # 2	0.232	17:26
Breath Volume	2.222	17:26
Air Blank	0.000	17:27
Control Test	0.079	17:27
Air Blank	0.000	17:28
Diagnostics Check	OK	17:28

Data Download Date/Time: 09/28/2017 11:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THIGPEN,J,T
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: VANHOF,J
ARRESTING OFFICER AGENCY: Atlantic Beach PD
LAST AGENCY INSPECTION: 09/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 18:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:20

SUBJECT'S LAST NAME: BOYD

DATE/TIME OF TEST: 09/15/2017 20:47:09

Diagnostics Check	OK	20:40
Air Blank	0.000	20:41
Control Test	0.079	20:41
Air Blank	0.000	20:42
Subject Sample # 1	0.230	20:42
Breath Volume	3.835	20:42
Air Blank	0.000	20:43
Air Blank	0.000	20:45
Subject Sample # 2	0.228	20:45
Breath Volume	3.621	20:45
Air Blank	0.000	20:46
Control Test	0.078	20:46
Air Blank	0.000	20:47
Diagnostics Check	OK	20:47

Data Download Date/Time: 09/28/2017 11:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THIGPEN,J,T
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BYRNE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 09/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:08

SUBJECT'S LAST NAME: O RIORDAN

DATE/TIME OF TEST: 09/15/2017 21:35:11

Diagnostics Check	OK	21:28
Air Blank	0.000	21:28
Control Test	0.079	21:29
Air Blank	0.000	21:29
Subject Sample # 1	0.218	21:30
Breath Volume	1.281	21:30
Air Blank	0.000	21:31
Air Blank	0.000	21:33
Subject Sample # 2	0.214	21:33
Breath Volume	1.175	21:33
Air Blank	0.000	21:34
Control Test	0.078	21:34
Air Blank	0.000	21:35
Diagnostics Check	OK	21:35

Data Download Date/Time: 09/28/2017 11:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MURRIELL,E,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: GARCIA,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 09/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 16:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:28

SUBJECT'S LAST NAME: DOHERTY

DATE/TIME OF TEST: 09/20/2017 17:55:30

Diagnostics Check	OK	17:49
Air Blank	0.000	17:49
Control Test	0.080	17:49
Air Blank	0.000	17:50
Subject Sample # 1	0.000	17:51
Breath Volume	1.691	17:51
Air Blank	0.000	17:51
Air Blank	0.000	17:53
Subject Sample # 2	0.000	17:54
Breath Volume	1.113	17:54
Air Blank	0.000	17:54
Control Test	0.080	17:54
Air Blank	0.000	17:55
Diagnostics Check	OK	17:55

Data Download Date/Time: 09/28/2017 11:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MURRIELL,E,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FISETTE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 09/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 17:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:02

SUBJECT'S LAST NAME: GIBSON

DATE/TIME OF TEST: 09/20/2017 20:27:45

Diagnostics Check	OK	20:22	
Air Blank	0.000	20:23	
Control Test	0.080	20:23	
Air Blank	0.000	20:24	
Subject Sample	# 1	20:26	Subject Test Refused
Breath Volume	0.757	20:26	
Air Blank	0.000	20:26	
Control Test	0.079	20:27	
Air Blank	0.000	20:27	
Diagnostics Check	OK	20:27	

Data Download Date/Time: 09/28/2017 11:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MURRIELL,E,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHITE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 09/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:09

SUBJECT'S LAST NAME: REWIS

DATE/TIME OF TEST: 09/20/2017 22:33:09

Diagnostics Check	OK	22:29	
Air Blank	0.000	22:30	
Control Test	0.078	22:30	
Air Blank	0.000	22:31	
Subject Sample	# 1	22:31	Subject Test Refused
Breath Volume	0.000	22:31	
Air Blank	0.000	22:32	
Control Test	0.080	22:32	
Air Blank	0.000	22:33	
Diagnostics Check	OK	22:33	

Data Download Date/Time: 09/28/2017 11:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THIGPEN,J,T
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BRYANT,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 09/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 16:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:24

SUBJECT'S LAST NAME: KIRKLAND

DATE/TIME OF TEST: 09/23/2017 18:51:10

Diagnostics Check	OK	18:44
Air Blank	0.000	18:45
Control Test	0.079	18:45
Air Blank	0.000	18:46
Subject Sample # 1	0.000	18:46
Breath Volume	2.878	18:46
Air Blank	0.000	18:47
Air Blank	0.000	18:49
Subject Sample # 2	0.000	18:49
Breath Volume	3.605	18:49
Air Blank	0.000	18:50
Control Test	0.080	18:50
Air Blank	0.000	18:51
Diagnostics Check	OK	18:51

Data Download Date/Time: 09/28/2017 11:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THIGPEN,J,T
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MARRERO,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 09/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 17:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:17

SUBJECT'S LAST NAME: BUTLER

DATE/TIME OF TEST: 09/23/2017 20:51:28

Diagnostics Check	OK	20:37	
Air Blank	0.000	20:38	
Control Test	0.080	20:38	
Air Blank	0.000	20:39	
Subject Sample # 1	0.000	20:41	
Breath Volume	1.132	20:41	
Air Blank	0.000	20:41	
Air Blank	0.000	20:43	
Subject Sample # 2	0.000	20:46	Volume Not Met
Breath Volume	0.613	20:46	
Air Blank	0.000	20:47	
Air Blank	0.000	20:49	
Subject Sample # 3	0.000	20:50	
Breath Volume	1.402	20:50	
Air Blank	0.000	20:50	
Control Test	0.079	20:50	
Air Blank	0.000	20:51	
Diagnostics Check	OK	20:51	

Data Download Date/Time: 09/28/2017 11:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THIGPEN,J,T
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HEALY,G
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 09/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:17

SUBJECT'S LAST NAME: KELLY

DATE/TIME OF TEST: 09/23/2017 22:40:41

Diagnostics Check	OK	22:37	
Air Blank	0.000	22:38	
Control Test	0.079	22:38	
Air Blank	0.000	22:39	
Subject Sample	# 1	22:39	Subject Test Refused
Breath Volume	0.000	22:39	
Air Blank	0.000	22:39	
Control Test	0.080	22:40	
Air Blank	0.000	22:40	
Diagnostics Check	OK	22:40	

Data Download Date/Time: 09/28/2017 11:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THIGPEN,J,T
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: KEEN,C
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 09/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 18:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:40

SUBJECT'S LAST NAME: MILLIKEN

DATE/TIME OF TEST: 09/24/2017 21:07:55

Diagnostics Check	OK	21:01
Air Blank	0.000	21:01
Control Test	0.079	21:02
Air Blank	0.000	21:02
Subject Sample # 1	0.219	21:03
Breath Volume	1.296	21:03
Air Blank	0.000	21:03
Air Blank	0.000	21:05
Subject Sample # 2	0.233	21:06
Breath Volume	2.027	21:06
Air Blank	0.000	21:07
Control Test	0.080	21:07
Air Blank	0.000	21:07
Diagnostics Check	OK	21:07

Data Download Date/Time: 09/28/2017 11:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THIGPEN,J,T
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MARRERO,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 09/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: TOUSENT

DATE/TIME OF TEST: 09/24/2017 22:24:04

Diagnostics Check	OK	22:21	
Air Blank	0.000	22:21	
Control Test	0.079	22:22	
Air Blank	0.000	22:22	
Subject Sample	# 1	22:22	Subject Test Refused
Breath Volume	0.000	22:22	
Air Blank	0.000	22:23	
Control Test	0.079	22:23	
Air Blank	0.000	22:23	
Diagnostics Check	OK	22:24	

Data Download Date/Time: 09/28/2017 11:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THIGPEN,J,T
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: STANLEY,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 09/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:51

SUBJECT'S LAST NAME: GATTO

DATE/TIME OF TEST: 09/26/2017 22:17:37

Diagnostics Check	OK	22:11
Air Blank	0.000	22:11
Control Test	0.080	22:12
Air Blank	0.000	22:12
Subject Sample # 1	0.027	22:13
Breath Volume	2.679	22:13
Air Blank	0.000	22:13
Air Blank	0.000	22:15
Subject Sample # 2	0.028	22:16
Breath Volume	2.089	22:16
Air Blank	0.000	22:16
Control Test	0.080	22:17
Air Blank	0.000	22:17
Diagnostics Check	OK	22:17

Data Download Date/Time: 09/28/2017 11:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,A.B.,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FILBERT,E
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 09/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 17:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:18

SUBJECT'S LAST NAME: FITZROY

DATE/TIME OF TEST: 09/27/2017 21:44:49

Diagnostics Check	OK	21:41	
Air Blank	0.000	21:42	
Control Test	0.079	21:42	
Air Blank	0.000	21:43	
Subject Sample	# 1	21:43	Subject Test Refused
Breath Volume	0.000	21:43	
Air Blank	0.000	21:43	
Control Test	0.079	21:44	
Air Blank	0.000	21:44	
Diagnostics Check	OK	21:44	

Data Download Date/Time: 09/28/2017 11:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,EUGENE,N
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: GARCIA,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 09/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:44

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 09/29/2017 22:10:57

Diagnostics Check	OK	22:04
Air Blank	0.000	22:05
Control Test	0.080	22:05
Air Blank	0.000	22:06
Subject Sample # 1	0.104	22:06
Breath Volume	1.414	22:06
Air Blank	0.000	22:07
Air Blank	0.000	22:09
Subject Sample # 2	0.104	22:09
Breath Volume	2.222	22:09
Air Blank	0.000	22:10
Control Test	0.079	22:10
Air Blank	0.000	22:10
Diagnostics Check	OK	22:10

Data Download Date/Time: 10/12/2017 12:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,J,T
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CHOWDHURY,P
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 09/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 15:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:50

SUBJECT'S LAST NAME: WOOTEN

DATE/TIME OF TEST: 09/30/2017 16:17:46

Diagnostics Check	OK	16:11
Air Blank	0.000	16:11
Control Test	0.080	16:12
Air Blank	0.000	16:12
Subject Sample # 1	0.000	16:13
Breath Volume	1.902	16:13
Air Blank	0.000	16:13
Air Blank	0.000	16:15
Subject Sample # 2	0.000	16:16
Breath Volume	2.050	16:16
Air Blank	0.000	16:16
Control Test	0.079	16:17
Air Blank	0.000	16:17
Diagnostics Check	OK	16:17

Data Download Date/Time: 10/12/2017 12:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,J,T
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BRYANT,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 09/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: HARTUNG

DATE/TIME OF TEST: 09/30/2017 21:43:51

Diagnostics Check	OK	21:41	
Air Blank	0.000	21:41	
Control Test	0.080	21:41	
Air Blank	0.000	21:42	
Subject Sample	# 1	21:42	Subject Test Refused
Breath Volume	0.000	21:42	
Air Blank	0.000	21:42	
Control Test	0.079	21:43	
Air Blank	0.000	21:43	
Diagnostics Check	OK	21:43	

Data Download Date/Time: 10/12/2017 12:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: STANLEY,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:57

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 09/02/2017 01:24:30

Diagnostics Check	OK	01:18
Air Blank	0.000	01:18
Control Test	0.079	01:18
Air Blank	0.000	01:19
Subject Sample # 1	0.138	01:19
Breath Volume	2.515	01:19
Air Blank	0.000	01:20
Air Blank	0.000	01:22
Subject Sample # 2	0.158	01:22
Breath Volume	1.585	01:22
Air Blank	0.000	01:23
Control Test	0.078	01:23
Air Blank	0.000	01:24
Diagnostics Check	OK	01:24

Data Download Date/Time: 09/12/2017 12:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY, SHANON, E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BRYANT, R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:26

SUBJECT'S LAST NAME: SNELL

DATE/TIME OF TEST: 09/06/2017 23:52:59

Diagnostics Check	OK	23:46
Air Blank	0.000	23:47
Control Test	0.078	23:47
Air Blank	0.000	23:48
Subject Sample # 1	0.195	23:48
Breath Volume	1.417	23:48
Air Blank	0.000	23:49
Air Blank	0.000	23:51
Subject Sample # 2	0.211	23:51
Breath Volume	1.808	23:51
Air Blank	0.000	23:52
Control Test	0.079	23:52
Air Blank	0.000	23:52
Diagnostics Check	OK	23:52

Data Download Date/Time: 09/12/2017 12:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY,SHANON,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: KEEN,C
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 08/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:36

SUBJECT'S LAST NAME: LONG

DATE/TIME OF TEST: 09/07/2017 05:03:01

Diagnostics Check	OK	04:56
Air Blank	0.000	04:57
Control Test	0.079	04:57
Air Blank	0.000	04:57
Subject Sample # 1	0.121	04:58
Breath Volume	1.406	04:58
Air Blank	0.000	04:59
Air Blank	0.000	05:01
Subject Sample # 2	0.108	05:01
Breath Volume	1.511	05:01
Air Blank	0.000	05:02
Control Test	0.080	05:02
Air Blank	0.000	05:02
Diagnostics Check	OK	05:03

Data Download Date/Time: 09/12/2017 12:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FILBERT,E
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 08/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: PAOLONE

DATE/TIME OF TEST: 09/08/2017 03:37:48

Diagnostics Check	OK	03:30
Air Blank	0.000	03:31
Control Test	0.079	03:31
Air Blank	0.000	03:32
Subject Sample # 1	0.125	03:33
Breath Volume	1.148	03:33
Air Blank	0.000	03:33
Air Blank	0.000	03:35
Subject Sample # 2	0.119	03:36
Breath Volume	1.640	03:36
Air Blank	0.000	03:36
Control Test	0.078	03:37
Air Blank	0.000	03:37
Diagnostics Check	OK	03:37

Data Download Date/Time: 09/12/2017 12:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POWELL,EDWARD,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MCKEITHEN,J
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 08/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:16

SUBJECT'S LAST NAME: REED

DATE/TIME OF TEST: 09/08/2017 04:40:33

Diagnostics Check	OK	04:37	
Air Blank	0.000	04:38	
Control Test	0.079	04:38	
Air Blank	0.000	04:38	
Subject Sample	# 1	04:39	Subject Test Refused
Breath Volume	0.000	04:39	
Air Blank	0.000	04:39	
Control Test	0.080	04:39	
Air Blank	0.000	04:40	
Diagnostics Check	OK	04:40	

Data Download Date/Time: 09/12/2017 12:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEE,JAMAL,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BRETT,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:31

SUBJECT'S LAST NAME: TOPIK

DATE/TIME OF TEST: 09/08/2017 07:55:09

Diagnostics Check	OK	07:51	
Air Blank	0.000	07:52	
Control Test	0.078	07:52	
Air Blank	0.000	07:53	
Subject Sample	# 1	07:53	Subject Test Refused
Breath Volume	0.000	07:53	
Air Blank	0.000	07:54	
Control Test	0.079	07:54	
Air Blank	0.000	07:55	
Diagnostics Check	OK	07:55	

Data Download Date/Time: 09/12/2017 12:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: KEEN,C
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 08/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:56

SUBJECT'S LAST NAME: PATANIA

DATE/TIME OF TEST: 09/08/2017 23:19:43

Diagnostics Check	OK	23:16	
Air Blank	0.000	23:17	
Control Test	0.080	23:17	
Air Blank	0.000	23:18	
Subject Sample	# 1	23:18	Subject Test Refused
Breath Volume	0.000	23:18	
Air Blank	0.000	23:18	
Control Test	0.079	23:19	
Air Blank	0.000	23:19	
Diagnostics Check	OK	23:19	

Data Download Date/Time: 09/12/2017 12:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOSTON,DONALD,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: PROTONOTARIOUS,N
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 05:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:57

SUBJECT'S LAST NAME: NANCE

DATE/TIME OF TEST: 09/09/2017 09:24:58

Diagnostics Check	OK	09:18
Air Blank	0.000	09:18
Control Test	0.079	09:18
Air Blank	0.000	09:19
Subject Sample # 1	0.135	09:20
Breath Volume	2.253	09:20
Air Blank	0.000	09:20
Air Blank	0.000	09:22
Subject Sample # 2	0.146	09:23
Breath Volume	2.183	09:23
Air Blank	0.000	09:24
Control Test	0.079	09:24
Air Blank	0.000	09:24
Diagnostics Check	OK	09:24

Data Download Date/Time: 09/12/2017 12:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POWELL,EDWARD,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DOUGHFMAN,D
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:53

SUBJECT'S LAST NAME: GRIGGS

DATE/TIME OF TEST: 09/10/2017 00:19:53

Diagnostics Check	OK	00:13
Air Blank	0.000	00:13
Control Test	0.079	00:14
Air Blank	0.000	00:14
Subject Sample # 1	0.151	00:15
Breath Volume	2.046	00:15
Air Blank	0.000	00:15
Air Blank	0.000	00:17
Subject Sample # 2	0.146	00:18
Breath Volume	2.187	00:18
Air Blank	0.000	00:18
Control Test	0.079	00:19
Air Blank	0.000	00:19
Diagnostics Check	OK	00:19

Data Download Date/Time: 09/12/2017 12:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MORGAN,J
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 08/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 09/10/2017 03:37:23

Diagnostics Check	OK	03:30
Air Blank	0.000	03:31
Control Test	0.079	03:31
Air Blank	0.000	03:31
Subject Sample # 1	0.105	03:32
Breath Volume	1.636	03:32
Air Blank	0.000	03:33
Air Blank	0.000	03:35
Subject Sample # 2	0.104	03:35
Breath Volume	2.527	03:35
Air Blank	0.000	03:36
Control Test	0.077	03:36
Air Blank	0.000	03:37
Diagnostics Check	OK	03:37

Data Download Date/Time: 09/12/2017 12:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POWELL,EDWARD,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HILLIER,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 05:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:34

SUBJECT'S LAST NAME: GARRETT

DATE/TIME OF TEST: 09/10/2017 07:01:44

Diagnostics Check	OK	06:54
Air Blank	0.000	06:54
Control Test	0.079	06:55
Air Blank	0.000	06:55
Subject Sample # 1	0.156	06:56
Breath Volume	1.785	06:56
Air Blank	0.000	06:57
Air Blank	0.000	06:59
Subject Sample # 2	0.163	07:00
Breath Volume	1.949	07:00
Air Blank	0.000	07:00
Control Test	0.079	07:01
Air Blank	0.000	07:01
Diagnostics Check	OK	07:01

Data Download Date/Time: 09/12/2017 12:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEE,JAMAL,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CARRION,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 09/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 09:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:25

SUBJECT'S LAST NAME: ROGUE

DATE/TIME OF TEST: 09/12/2017 12:54:48

Diagnostics Check	OK	12:47
Air Blank	0.000	12:47
Control Test	0.080	12:48
Air Blank	0.000	12:48
Subject Sample # 1	0.000	12:49
Breath Volume	4.476	12:49
Air Blank	0.000	12:50
Air Blank	0.000	12:52
Subject Sample # 2	0.000	12:53
Breath Volume	4.019	12:53
Air Blank	0.000	12:53
Control Test	0.080	12:54
Air Blank	0.000	12:54
Diagnostics Check	OK	12:54

Data Download Date/Time: 09/28/2017 11:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY,SHANON,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MOELLER,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 09/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: RUSHITAJ

DATE/TIME OF TEST: 09/17/2017 02:56:53

Diagnostics Check	OK	02:50
Air Blank	0.000	02:51
Control Test	0.079	02:51
Air Blank	0.000	02:51
Subject Sample # 1	0.062	02:52
Breath Volume	2.250	02:52
Air Blank	0.000	02:52
Air Blank	0.000	02:54
Subject Sample # 2	0.062	02:55
Breath Volume	2.007	02:55
Air Blank	0.000	02:55
Control Test	0.079	02:56
Air Blank	0.000	02:56
Diagnostics Check	OK	02:56

Data Download Date/Time: 09/28/2017 11:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY,SHANON,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CHERENA,F
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 09/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:48

SUBJECT'S LAST NAME: RAMALES

DATE/TIME OF TEST: 09/17/2017 05:16:14

Diagnostics Check	OK	05:08
Air Blank	0.000	05:09
Control Test	0.079	05:09
Air Blank	0.000	05:10
Subject Sample # 1	0.153	05:11
Breath Volume	2.816	05:11
Air Blank	0.000	05:12
Air Blank	0.000	05:14
Subject Sample # 2	0.151	05:14
Breath Volume	2.105	05:14
Air Blank	0.000	05:15
Control Test	0.078	05:15
Air Blank	0.000	05:16
Diagnostics Check	OK	05:16

Data Download Date/Time: 09/28/2017 11:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY,SHANON,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: GARCIA,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 09/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:46

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 09/18/2017 02:11:13

Diagnostics Check	OK	02:08	
Air Blank	0.000	02:08	
Control Test	0.079	02:09	
Air Blank	0.000	02:09	
Subject Sample	# 1	02:09	Subject Test Refused
Breath Volume	0.000	02:09	
Air Blank	0.000	02:10	
Control Test	0.079	02:10	
Air Blank	0.000	02:11	
Diagnostics Check	OK	02:11	

Data Download Date/Time: 09/28/2017 11:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY, SHANON, E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FILBERT, E
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 09/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:47

SUBJECT'S LAST NAME: EJMALI

DATE/TIME OF TEST: 09/18/2017 03:14:38

Diagnostics Check	OK	03:08
Air Blank	0.000	03:08
Control Test	0.080	03:08
Air Blank	0.000	03:09
Subject Sample # 1	0.141	03:10
Breath Volume	1.773	03:10
Air Blank	0.000	03:10
Air Blank	0.000	03:12
Subject Sample # 2	0.140	03:13
Breath Volume	1.871	03:13
Air Blank	0.000	03:13
Control Test	0.079	03:14
Air Blank	0.000	03:14
Diagnostics Check	OK	03:14

Data Download Date/Time: 09/28/2017 11:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY,SHANON,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: RODRIGUEZ,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 09/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:07

SUBJECT'S LAST NAME: ROMERO

DATE/TIME OF TEST: 09/18/2017 05:35:18

Diagnostics Check		OK	05:27
Air Blank		0.000	05:28
Control Test		0.080	05:28
Air Blank		0.000	05:29
Subject Sample	# 1	0.168	05:30
Breath Volume		1.519	05:30
Air Blank		0.000	05:30
Air Blank		0.000	05:32
Subject Sample	# 2	0.167	05:33
Breath Volume		1.121	05:33
Air Blank		0.000	05:34
Control Test		0.077	05:34
Air Blank		0.000	05:35
Diagnostics Check		OK	05:35

Data Download Date/Time: 09/28/2017 11:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY,BERRY,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FISETTE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 09/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:23

SUBJECT'S LAST NAME: FULMER

DATE/TIME OF TEST: 09/18/2017 23:50:13

Diagnostics Check	OK	23:43
Air Blank	0.000	23:44
Control Test	0.078	23:44
Air Blank	0.000	23:45
Subject Sample # 1	0.064	23:45
Breath Volume	2.000	23:45
Air Blank	0.000	23:46
Air Blank	0.000	23:48
Subject Sample # 2	0.063	23:48
Breath Volume	1.679	23:48
Air Blank	0.000	23:49
Control Test	0.079	23:49
Air Blank	0.000	23:50
Diagnostics Check	OK	23:50

Data Download Date/Time: 09/28/2017 11:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FARLEY,J
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 09/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:28

SUBJECT'S LAST NAME: EVANS

DATE/TIME OF TEST: 09/22/2017 01:55:11

Diagnostics Check	OK	01:48
Air Blank	0.000	01:48
Control Test	0.079	01:48
Air Blank	0.000	01:49
Subject Sample # 1	0.193	01:50
Breath Volume	1.453	01:50
Air Blank	0.000	01:50
Air Blank	0.000	01:52
Subject Sample # 2	0.192	01:53
Breath Volume	1.218	01:53
Air Blank	0.000	01:54
Control Test	0.078	01:54
Air Blank	0.000	01:55
Diagnostics Check	OK	01:55

Data Download Date/Time: 09/28/2017 11:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY,SHANON,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MARRERO,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 09/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: LONG

DATE/TIME OF TEST: 09/22/2017 03:13:51

Diagnostics Check	OK	03:11	
Air Blank	0.000	03:11	
Control Test	0.079	03:11	
Air Blank	0.000	03:12	
Subject Sample	# 1	03:12	Subject Test Refused
Breath Volume	0.000	03:12	
Air Blank	0.000	03:12	
Control Test	0.080	03:13	
Air Blank	0.000	03:13	
Diagnostics Check	OK	03:13	

Data Download Date/Time: 09/28/2017 11:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BYRNE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 09/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:03

SUBJECT'S LAST NAME: SINGLETON

DATE/TIME OF TEST: 09/23/2017 23:27:38

Diagnostics Check	OK	23:24	
Air Blank	0.000	23:25	
Control Test	0.080	23:25	
Air Blank	0.000	23:26	
Subject Sample	# 1	23:26	Subject Test Refused
Breath Volume	0.000	23:26	
Air Blank	0.000	23:26	
Control Test	0.079	23:27	
Air Blank	0.000	23:27	
Diagnostics Check	OK	23:27	

Data Download Date/Time: 09/28/2017 11:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: KEEN,C
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 09/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:03

SUBJECT'S LAST NAME: CHRISTOPHER

DATE/TIME OF TEST: 09/24/2017 04:30:03

Diagnostics Check		OK	04:23
Air Blank		0.000	04:24
Control Test		0.079	04:24
Air Blank		0.000	04:24
Subject Sample	# 1	0.145	04:25
Breath Volume		2.261	04:25
Air Blank		0.000	04:26
Air Blank		0.000	04:28
Subject Sample	# 2	0.150	04:28
Breath Volume		2.019	04:28
Air Blank		0.000	04:29
Control Test		0.079	04:29
Air Blank		0.000	04:29
Diagnostics Check		OK	04:30

Data Download Date/Time: 09/28/2017 11:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BYRNE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 09/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:49

SUBJECT'S LAST NAME: HAYES

DATE/TIME OF TEST: 09/24/2017 23:15:55

Diagnostics Check	OK	23:09
Air Blank	0.000	23:09
Control Test	0.079	23:09
Air Blank	0.000	23:10
Subject Sample # 1	0.156	23:10
Breath Volume	1.671	23:10
Air Blank	0.000	23:11
Air Blank	0.000	23:13
Subject Sample # 2	0.147	23:14
Breath Volume	1.382	23:14
Air Blank	0.000	23:15
Control Test	0.078	23:15
Air Blank	0.000	23:15
Diagnostics Check	OK	23:15

Data Download Date/Time: 09/28/2017 11:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BRYANT,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 09/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:47

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 09/25/2017 01:14:16

Diagnostics Check		OK	01:08
Air Blank		0.000	01:08
Control Test		0.079	01:08
Air Blank		0.000	01:09
Subject Sample	# 1	0.152	01:09
Breath Volume		2.976	01:09
Air Blank		0.000	01:10
Air Blank		0.000	01:12
Subject Sample	# 2	0.156	01:12
Breath Volume		1.902	01:12
Air Blank		0.000	01:13
Control Test		0.079	01:13
Air Blank		0.000	01:14
Diagnostics Check		OK	01:14

Data Download Date/Time: 09/28/2017 11:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DYER,W
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 09/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: GARCIA TORRES

DATE/TIME OF TEST: 09/25/2017 03:38:20

Diagnostics Check	OK	03:35	
Air Blank	0.000	03:35	
Control Test	0.079	03:36	
Air Blank	0.000	03:36	
Subject Sample	# 1	03:36	Subject Test Refused
Breath Volume	0.000	03:36	
Air Blank	0.000	03:37	
Control Test	0.080	03:37	
Air Blank	0.000	03:38	
Diagnostics Check	OK	03:38	

Data Download Date/Time: 09/28/2017 11:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FILBERT,E
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 09/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:54

SUBJECT'S LAST NAME: RAMOS-SANDOVAL

DATE/TIME OF TEST: 09/26/2017 02:21:09

Diagnostics Check	OK	02:15
Air Blank	0.000	02:15
Control Test	0.080	02:15
Air Blank	0.000	02:16
Subject Sample # 1	0.000	02:16
Breath Volume	1.574	02:16
Air Blank	0.000	02:17
Air Blank	0.000	02:19
Subject Sample # 2	0.000	02:19
Breath Volume	2.632	02:19
Air Blank	0.000	02:20
Control Test	0.080	02:20
Air Blank	0.000	02:21
Diagnostics Check	OK	02:21

Data Download Date/Time: 09/28/2017 11:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BRYANT,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 09/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:07

SUBJECT'S LAST NAME: GODINEZ CARRILLO

DATE/TIME OF TEST: 09/26/2017 03:34:33

Diagnostics Check	OK	03:27
Air Blank	0.000	03:28
Control Test	0.079	03:28
Air Blank	0.000	03:29
Subject Sample # 1	0.203	03:30
Breath Volume	2.128	03:30
Air Blank	0.000	03:30
Air Blank	0.000	03:32
Subject Sample # 2	0.201	03:32
Breath Volume	1.562	03:32
Air Blank	0.000	03:33
Control Test	0.078	03:33
Air Blank	0.000	03:34
Diagnostics Check	OK	03:34

Data Download Date/Time: 09/28/2017 11:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPICER,LORETHA,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JOHNSON,G
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 09/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 08:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 12:43

SUBJECT'S LAST NAME: SUMPTER

DATE/TIME OF TEST: 09/26/2017 13:10:08

Diagnostics Check	OK	13:03
Air Blank	0.000	13:04
Control Test	0.080	13:04
Air Blank	0.000	13:05
Subject Sample # 1	0.115	13:05
Breath Volume	2.085	13:05
Air Blank	0.000	13:06
Air Blank	0.000	13:08
Subject Sample # 2	0.113	13:08
Breath Volume	2.320	13:08
Air Blank	0.000	13:09
Control Test	0.079	13:09
Air Blank	0.000	13:10
Diagnostics Check	OK	13:10

Data Download Date/Time: 09/28/2017 11:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY,SHANON,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CAMACHO,K
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 09/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:53

SUBJECT'S LAST NAME: MCEADY

DATE/TIME OF TEST: 09/28/2017 02:20:21

Diagnostics Check	OK	02:13
Air Blank	0.000	02:13
Control Test	0.079	02:14
Air Blank	0.000	02:14
Subject Sample # 1	0.160	02:15
Breath Volume	3.191	02:15
Air Blank	0.000	02:16
Air Blank	0.000	02:17
Subject Sample # 2	0.161	02:18
Breath Volume	3.332	02:18
Air Blank	0.000	02:19
Control Test	0.078	02:19
Air Blank	0.000	02:20
Diagnostics Check	OK	02:20

Data Download Date/Time: 09/28/2017 11:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY, SHANON, E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FISETTE, J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 09/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:54

SUBJECT'S LAST NAME: EYRE

DATE/TIME OF TEST: 09/29/2017 01:21:51

Diagnostics Check		OK	01:15
Air Blank		0.000	01:15
Control Test		0.080	01:15
Air Blank		0.000	01:16
Subject Sample	# 1	0.212	01:17
Breath Volume		2.738	01:17
Air Blank		0.000	01:18
Air Blank		0.000	01:19
Subject Sample	# 2	0.210	01:20
Breath Volume		2.929	01:20
Air Blank		0.000	01:20
Control Test		0.078	01:21
Air Blank		0.000	01:21
Diagnostics Check		OK	01:21

Data Download Date/Time: 10/12/2017 12:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY, SHANON, E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FARLEY, J
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 09/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:43

SUBJECT'S LAST NAME: ROUNTREE

DATE/TIME OF TEST: 09/29/2017 03:10:00

Diagnostics Check		OK	03:04
Air Blank		0.000	03:04
Control Test		0.079	03:04
Air Blank		0.000	03:05
Subject Sample	# 1	0.135	03:05
Breath Volume		1.296	03:05
Air Blank		0.000	03:06
Air Blank		0.000	03:08
Subject Sample	# 2	0.130	03:08
Breath Volume		1.132	03:08
Air Blank		0.000	03:09
Control Test		0.079	03:09
Air Blank		0.000	03:09
Diagnostics Check		OK	03:10

Data Download Date/Time: 10/12/2017 12:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY,SHANON,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MOELLER,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 09/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:04

SUBJECT'S LAST NAME: ALVAREZ

DATE/TIME OF TEST: 09/29/2017 04:27:56

Diagnostics Check	OK	04:25	
Air Blank	0.000	04:25	
Control Test	0.080	04:25	
Air Blank	0.000	04:26	
Subject Sample	# 1	04:26	Subject Test Refused
Breath Volume	0.000	04:26	
Air Blank	0.000	04:27	
Control Test	0.079	04:27	
Air Blank	0.000	04:27	
Diagnostics Check	OK	04:27	

Data Download Date/Time: 10/12/2017 12:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY, SHANON, E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DOUGHFMAN, D
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 09/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:36

SUBJECT'S LAST NAME: QUINONES ACOSTA

DATE/TIME OF TEST: 09/29/2017 06:03:23

Diagnostics Check	OK	05:56
Air Blank	0.000	05:56
Control Test	0.079	05:57
Air Blank	0.000	05:57
Subject Sample # 1	0.115	05:58
Breath Volume	2.753	05:58
Air Blank	0.000	05:59
Air Blank	0.000	06:01
Subject Sample # 2	0.114	06:01
Breath Volume	2.589	06:01
Air Blank	0.000	06:02
Control Test	0.079	06:02
Air Blank	0.000	06:03
Diagnostics Check	OK	06:03

Data Download Date/Time: 10/12/2017 12:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POWELL,EDWARD,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: RODRIQUENZ,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 09/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:31

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 09/30/2017 02:57:43

Diagnostics Check	OK	02:54	
Air Blank	0.000	02:55	
Control Test	0.079	02:55	
Air Blank	0.000	02:56	
Subject Sample	# 1	02:56	Subject Test Refused
Breath Volume	0.000	02:56	
Air Blank	0.000	02:56	
Control Test	0.080	02:57	
Air Blank	0.000	02:57	
Diagnostics Check	OK	02:57	

Data Download Date/Time: 10/12/2017 12:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POWELL,EDWARD,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CAMACHO,K
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 09/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:34

SUBJECT'S LAST NAME: BLAKESLEE

DATE/TIME OF TEST: 09/30/2017 04:01:35

Diagnostics Check	OK	03:54
Air Blank	0.000	03:55
Control Test	0.079	03:55
Air Blank	0.000	03:56
Subject Sample # 1	0.114	03:56
Breath Volume	1.820	03:56
Air Blank	0.000	03:57
Air Blank	0.000	03:59
Subject Sample # 2	0.114	04:00
Breath Volume	1.835	04:00
Air Blank	0.000	04:00
Control Test	0.079	04:00
Air Blank	0.000	04:01
Diagnostics Check	OK	04:01

Data Download Date/Time: 10/12/2017 12:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001282

INSTRUMENT REGISTERED TO: UNIVERSITY OF FL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZARATE,JESSICA,L
BREATH TEST OPERATOR AGENCY: U of FI PD
ARRESTING OFFICER: ZARATE,JESSICA
ARRESTING OFFICER AGENCY: U of FI PD
LAST AGENCY INSPECTION: 08/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A2
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:44

SUBJECT'S LAST NAME: ROBERTS

DATE/TIME OF TEST: 09/12/2017 00:24:24

Diagnostics Check	OK	00:13	
Air Blank	0.000	00:13	
Control Test	0.078	00:13	
Air Blank	0.000	00:14	
Subject Sample	# 1	00:17	No Sample Provided
Breath Volume	0.000	00:17	
Air Blank	0.000	00:17	
Air Blank	0.000	00:19	
Subject Sample	# 2	00:23	No Sample Provided
Breath Volume	0.000	00:23	
Air Blank	0.000	00:23	
Control Test	0.080	00:23	
Air Blank	0.000	00:24	
Diagnostics Check	OK	00:24	

Data Download Date/Time: 09/29/2017 14:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001282

INSTRUMENT REGISTERED TO: UNIVERSITY OF FL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS JR,STEPHANIE,G
BREATH TEST OPERATOR AGENCY: U of FI PD
ARRESTING OFFICER: WILLIAMS JR,STEPHANIE
ARRESTING OFFICER AGENCY: U of FI PD
LAST AGENCY INSPECTION: 08/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A2
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 09/15/2017 02:48:53

Diagnostics Check		OK	02:38
Air Blank		0.000	02:38
Control Test		0.079	02:38
Air Blank		0.000	02:39
Subject Sample	# 1	0.236	02:40
Breath Volume		2.218	02:40
Air Blank		0.000	02:41
Air Blank		0.000	02:43
Subject Sample	# 2	0.205	02:43
Breath Volume		1.558	02:43
Air Blank		0.000	02:44
Air Blank		0.000	02:46
Subject Sample	# 3	0.235	02:47
Breath Volume		2.312	02:47
Air Blank		0.000	02:47
Control Test		0.078	02:48
Air Blank		0.000	02:48
Diagnostics Check		OK	02:48

Data Download Date/Time: 09/29/2017 14:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001282

INSTRUMENT REGISTERED TO: UNIVERSITY OF FL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZARATE,JESSICA,L
BREATH TEST OPERATOR AGENCY: U of FI PD
ARRESTING OFFICER: LEECH,EMMA
ARRESTING OFFICER AGENCY: U of FI PD
LAST AGENCY INSPECTION: 08/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A2
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:27

SUBJECT'S LAST NAME: HAY

DATE/TIME OF TEST: 09/16/2017 03:56:26

Diagnostics Check		OK	03:50
Air Blank		0.000	03:50
Control Test		0.078	03:50
Air Blank		0.000	03:51
Subject Sample	# 1	0.198	03:51
Breath Volume		2.417	03:51
Air Blank		0.000	03:52
Air Blank		0.000	03:54
Subject Sample	# 2	0.206	03:54
Breath Volume		2.675	03:54
Air Blank		0.000	03:55
Control Test		0.077	03:55
Air Blank		0.000	03:56
Diagnostics Check		OK	03:56

Data Download Date/Time: 09/29/2017 14:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001282

INSTRUMENT REGISTERED TO: UNIVERSITY OF FL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BECKHAM,MARCUS,W
BREATH TEST OPERATOR AGENCY: U of FI PD
ARRESTING OFFICER: ROBERTS,BRADLEY
ARRESTING OFFICER AGENCY: U of FI PD
LAST AGENCY INSPECTION: 08/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A2
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 16:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:14

SUBJECT'S LAST NAME: WINESETT

DATE/TIME OF TEST: 09/16/2017 16:42:21

Diagnostics Check	OK	16:35
Air Blank	0.000	16:36
Control Test	0.079	16:36
Air Blank	0.000	16:36
Subject Sample # 1	0.200	16:37
Breath Volume	1.683	16:37
Air Blank	0.000	16:38
Air Blank	0.000	16:39
Subject Sample # 2	0.197	16:40
Breath Volume	1.253	16:40
Air Blank	0.000	16:41
Control Test	0.079	16:41
Air Blank	0.000	16:42
Diagnostics Check	OK	16:42

Data Download Date/Time: 09/29/2017 14:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001282

INSTRUMENT REGISTERED TO: UNIVERSITY OF FL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHEFFIELD,KHRIS,D
BREATH TEST OPERATOR AGENCY: U of FI PD
ARRESTING OFFICER: ALEXANDER,COURTNEY
ARRESTING OFFICER AGENCY: U of FI PD
LAST AGENCY INSPECTION: 08/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A2
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:31

SUBJECT'S LAST NAME: SANCHEZ

DATE/TIME OF TEST: 09/17/2017 04:03:13

Diagnostics Check	OK	03:51	
Air Blank	0.000	03:52	
Control Test	0.078	03:52	
Air Blank	0.000	03:53	
Subject Sample	# 1	03:56	No Sample Provided
Breath Volume	0.000	03:56	
Air Blank	0.000	03:56	
Air Blank	0.000	03:58	
Subject Sample	# 2	04:01	No Sample Provided
Breath Volume	0.000	04:01	
Air Blank	0.000	04:02	
Control Test	0.078	04:02	
Air Blank	0.000	04:03	
Diagnostics Check	OK	04:03	

Data Download Date/Time: 09/29/2017 14:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001282

INSTRUMENT REGISTERED TO: UNIVERSITY OF FL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZARATE,JESSICA,L
BREATH TEST OPERATOR AGENCY: U of FI PD
ARRESTING OFFICER: ROGERS,MELISSA
ARRESTING OFFICER AGENCY: U of FI PD
LAST AGENCY INSPECTION: 08/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A2
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:44

SUBJECT'S LAST NAME: AYOROA GALINDO

DATE/TIME OF TEST: 09/17/2017 04:19:32

Diagnostics Check	OK	04:13
Air Blank	0.000	04:13
Control Test	0.079	04:13
Air Blank	0.000	04:14
Subject Sample # 1	0.182	04:14
Breath Volume	1.984	04:14
Air Blank	0.000	04:15
Air Blank	0.000	04:17
Subject Sample # 2	0.180	04:17
Breath Volume	3.343	04:17
Air Blank	0.000	04:18
Control Test	0.078	04:18
Air Blank	0.000	04:19
Diagnostics Check	OK	04:19

Data Download Date/Time: 09/29/2017 14:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001282

INSTRUMENT REGISTERED TO: UNIVERSITY OF FL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZARATE,JESSICA,L
BREATH TEST OPERATOR AGENCY: U of FI PD
ARRESTING OFFICER: ZARATE,JESSICA
ARRESTING OFFICER AGENCY: U of FI PD
LAST AGENCY INSPECTION: 08/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A2
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:32

SUBJECT'S LAST NAME: BREWER

DATE/TIME OF TEST: 09/22/2017 03:02:49

Diagnostics Check	OK	02:55
Air Blank	0.000	02:56
Control Test	0.079	02:56
Air Blank	0.000	02:56
Subject Sample # 1	0.095	02:58
Breath Volume	3.851	02:58
Air Blank	0.000	02:58
Air Blank	0.000	03:00
Subject Sample # 2	0.093	03:01
Breath Volume	3.765	03:01
Air Blank	0.000	03:01
Control Test	0.077	03:02
Air Blank	0.000	03:02
Diagnostics Check	OK	03:02

Data Download Date/Time: 09/29/2017 14:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001282

INSTRUMENT REGISTERED TO: UNIVERSITY OF FL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAIN II,KENNETH,A
BREATH TEST OPERATOR AGENCY: U of FI PD
ARRESTING OFFICER: NGUYEN,RICK
ARRESTING OFFICER AGENCY: U of FI PD
LAST AGENCY INSPECTION: 08/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A2
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:49

SUBJECT'S LAST NAME: CUADRADO CASTILLO

DATE/TIME OF TEST: 09/23/2017 01:22:47

Diagnostics Check	OK	01:16
Air Blank	0.000	01:16
Control Test	0.079	01:17
Air Blank	0.000	01:17
Subject Sample # 1	0.147	01:18
Breath Volume	1.164	01:18
Air Blank	0.000	01:18
Air Blank	0.000	01:20
Subject Sample # 2	0.145	01:21
Breath Volume	1.582	01:21
Air Blank	0.000	01:21
Control Test	0.079	01:22
Air Blank	0.000	01:22
Diagnostics Check	OK	01:22

Data Download Date/Time: 09/29/2017 14:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001285

INSTRUMENT REGISTERED TO: GILCHRIST COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FULTON,DARIN,M
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: FULTON,DARIN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 08/09/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 259585
DRY GAS EXPIRY DATE: 01/12/2019
VIOLATION CODE: DUI
ARREST TIME: 10:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: KARAMOLENGOS

DATE/TIME OF TEST: 09/03/2017 01:32:03

Diagnostics Check	OK	01:27	
Air Blank	0.000	01:27	
Control Test	0.081	01:27	
Air Blank	0.000	01:28	
Subject Sample	# 1	01:30	Subject Test Refused
Breath Volume	0.000	01:30	
Air Blank	0.000	01:31	
Control Test	0.080	01:31	
Air Blank	0.000	01:31	
Diagnostics Check	OK	01:32	

Data Download Date/Time: 09/24/2017 14:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001285

INSTRUMENT REGISTERED TO: GILCHRIST COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARAT,BRAD,L
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: MARAT,BRAD
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 08/09/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 259585
DRY GAS EXPIRY DATE: 01/12/2019
VIOLATION CODE: DUI
ARREST TIME: 04:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:08

SUBJECT'S LAST NAME: FLOYD

DATE/TIME OF TEST: 09/16/2017 05:35:17

Diagnostics Check	OK	05:28
Air Blank	0.000	05:29
Control Test	0.077	05:29
Air Blank	0.000	05:30
Subject Sample # 1	0.244	05:30
Breath Volume	4.035	05:30
Air Blank	0.000	05:31
Air Blank	0.000	05:33
Subject Sample # 2	0.243	05:33
Breath Volume	3.925	05:33
Air Blank	0.000	05:34
Control Test	0.079	05:34
Air Blank	0.000	05:35
Diagnostics Check	OK	05:35

Data Download Date/Time: 09/24/2017 14:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001287

INSTRUMENT REGISTERED TO: BAKER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMATT,JUSTIN,M
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: SATTERWHITE,JACOB
ARRESTING OFFICER AGENCY: Baker County SO
LAST AGENCY INSPECTION: 08/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:33

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 09/03/2017 00:01:09

Diagnostics Check	OK	23:54
Air Blank	0.000	23:54
Control Test	0.082	23:55
Air Blank	0.000	23:55
Subject Sample # 1	0.256	23:56
Breath Volume	1.343	23:56
Air Blank	0.000	23:56
Air Blank	0.000	23:58
Subject Sample # 2	0.262	23:59
Breath Volume	2.207	23:59
Air Blank	0.000	00:00
Control Test	0.081	00:00
Air Blank	0.000	00:01
Diagnostics Check	OK	00:01

Data Download Date/Time: 09/03/2017 15:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001287

INSTRUMENT REGISTERED TO: BAKER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,DANNY,L
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MYERS,DANNY
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 08/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:45

SUBJECT'S LAST NAME: DRABIK

DATE/TIME OF TEST: 09/21/2017 21:18:49

Diagnostics Check		OK	21:12
Air Blank		0.000	21:13
Control Test		0.081	21:13
Air Blank		0.000	21:13
Subject Sample	# 1	0.186	21:14
Breath Volume		2.691	21:14
Air Blank		0.000	21:15
Air Blank		0.000	21:16
Subject Sample	# 2	0.186	21:17
Breath Volume		2.167	21:17
Air Blank		0.000	21:17
Control Test		0.080	21:18
Air Blank		0.000	21:18
Diagnostics Check		OK	21:18

Data Download Date/Time: 10/02/2017 01:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001287

INSTRUMENT REGISTERED TO: BAKER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARLEY,JOSEPH,E
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: FARLEY,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 08/31/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:31

SUBJECT'S LAST NAME: LLOYD

DATE/TIME OF TEST: 09/24/2017 00:58:05

Diagnostics Check	OK	00:51
Air Blank	0.000	00:52
Control Test	0.082	00:52
Air Blank	0.000	00:53
Subject Sample # 1	0.082	00:53
Breath Volume	3.195	00:53
Air Blank	0.000	00:54
Air Blank	0.000	00:55
Subject Sample # 2	0.080	00:56
Breath Volume	3.328	00:56
Air Blank	0.000	00:57
Control Test	0.081	00:57
Air Blank	0.000	00:58
Diagnostics Check	OK	00:58

Data Download Date/Time: 10/02/2017 01:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001287

INSTRUMENT REGISTERED TO: BAKER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARLEY,JOSEPH,E
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: FARLEY,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 08/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:42

SUBJECT'S LAST NAME: GASKINS

DATE/TIME OF TEST: 09/24/2017 04:09:12

Diagnostics Check	OK	04:02
Air Blank	0.000	04:03
Control Test	0.081	04:03
Air Blank	0.000	04:04
Subject Sample # 1	0.239	04:04
Breath Volume	2.531	04:04
Air Blank	0.000	04:05
Air Blank	0.000	04:07
Subject Sample # 2	0.258	04:07
Breath Volume	2.871	04:07
Air Blank	0.000	04:08
Control Test	0.081	04:08
Air Blank	0.000	04:09
Diagnostics Check	OK	04:09

Data Download Date/Time: 10/02/2017 01:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001289

INSTRUMENT REGISTERED TO: UNION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HILLIARD,KARLA
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: BOATRIGHT,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 08/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1824329
DRY GAS EXPIRY DATE: 03/09/2018
VIOLATION CODE: DUI
ARREST TIME: 01:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: YOUNG

DATE/TIME OF TEST: 09/03/2017 02:24:21

Diagnostics Check	OK	02:15	
Air Blank	0.000	02:16	
Control Test	0.081	02:16	
Air Blank	0.000	02:17	
Subject Sample	# 1	02:20	No Sample Provided
Breath Volume	0.000	02:20	
Air Blank	0.000	02:20	
Air Blank	0.000	02:22	
Subject Sample	# 2	02:23	RFI Detect
Breath Volume	0.000	02:23	
Air Blank	0.000	02:23	
Control Test	0.081	02:23	
Air Blank	0.000	02:24	
Diagnostics Check	OK	02:24	

Data Download Date/Time: 10/10/2017 16:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001289

INSTRUMENT REGISTERED TO: UNION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HILLIARD,KARLA
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: BOATRIGHT,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 08/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1824329
DRY GAS EXPIRY DATE: 03/09/2018
VIOLATION CODE: DUI
ARREST TIME: 01:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: YOUNG

DATE/TIME OF TEST: 09/03/2017 02:32:20

Diagnostics Check	OK	02:29	
Air Blank	0.000	02:30	
Control Test	0.080	02:30	
Air Blank	0.000	02:30	
Subject Sample	# 1	02:30	Subject Test Refused
Breath Volume	0.000	02:30	
Air Blank	0.000	02:31	
Control Test	0.081	02:31	
Air Blank	0.000	02:32	
Diagnostics Check	OK	02:32	

Data Download Date/Time: 10/10/2017 16:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001289

INSTRUMENT REGISTERED TO: UNION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JANKOWSKI,JAMES,F
BREATH TEST OPERATOR AGENCY: Union County SO
ARRESTING OFFICER: JANKOWSKI,JAMES
ARRESTING OFFICER AGENCY: Union County SO
LAST AGENCY INSPECTION: 08/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1824329
DRY GAS EXPIRY DATE: 03/09/2018
VIOLATION CODE: DUI
ARREST TIME: 19:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:22

SUBJECT'S LAST NAME: MILEY

DATE/TIME OF TEST: 09/16/2017 20:49:23

Diagnostics Check	OK	20:43	
Air Blank	0.000	20:44	
Control Test	0.079	20:44	
Air Blank	0.000	20:45	
Subject Sample	# 1	20:47	Subject Test Refused
Breath Volume	1.085	20:47	
Air Blank	0.000	20:48	
Control Test	0.081	20:48	
Air Blank	0.000	20:49	
Diagnostics Check	OK	20:49	

Data Download Date/Time: 10/10/2017 16:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001290

INSTRUMENT REGISTERED TO: BRADFORD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARLEY,JOSEPH,E
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: FARLEY,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 08/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG619605
DRY GAS EXPIRY DATE: 07/14/2018
VIOLATION CODE: DUI
ARREST TIME: 00:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:46

SUBJECT'S LAST NAME: KEARSE

DATE/TIME OF TEST: 09/13/2017 01:24:49

Diagnostics Check		OK	01:10	
Air Blank		0.000	01:11	
Control Test		0.079	01:11	
Air Blank		0.000	01:12	
Subject Sample	# 1	0.155	01:15	Volume Not Met
Breath Volume		0.101	01:15	
Air Blank		0.000	01:15	
Air Blank		0.000	01:17	
Subject Sample	# 2	0.163	01:19	
Breath Volume		2.433	01:19	
Air Blank		0.000	01:19	
Air Blank		0.000	01:21	
Subject Sample	# 3	0.164	01:23	
Breath Volume		1.644	01:23	
Air Blank		0.000	01:23	
Control Test		0.079	01:24	
Air Blank		0.000	01:24	
Diagnostics Check		OK	01:24	

Data Download Date/Time: 09/24/2017 08:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001290

INSTRUMENT REGISTERED TO: BRADFORD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORRELL,JOSEPH,B
BREATH TEST OPERATOR AGENCY: Bradford County SO
ARRESTING OFFICER: MORRELL,JOSEPH
ARRESTING OFFICER AGENCY: Bradford County SO
LAST AGENCY INSPECTION: 08/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG619605
DRY GAS EXPIRY DATE: 07/14/2018
VIOLATION CODE: DUI
ARREST TIME: 20:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:34

SUBJECT'S LAST NAME: BIDDY

DATE/TIME OF TEST: 09/16/2017 20:57:50

Diagnostics Check	OK	20:54	
Air Blank	0.000	20:55	
Control Test	0.077	20:55	
Air Blank	0.000	20:56	
Subject Sample	# 1	20:56	Subject Test Refused
Breath Volume	0.000	20:56	
Air Blank	0.000	20:56	
Control Test	0.078	20:57	
Air Blank	0.000	20:57	
Diagnostics Check	OK	20:57	

Data Download Date/Time: 09/24/2017 08:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001292

INSTRUMENT REGISTERED TO: LEVY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LESHER, TIFFANY, A
BREATH TEST OPERATOR AGENCY: Levy County SO
ARRESTING OFFICER: SULLIVAN, RYAN
ARRESTING OFFICER AGENCY: Levy County SO
LAST AGENCY INSPECTION: 08/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG619605
DRY GAS EXPIRY DATE: 07/14/2018
VIOLATION CODE: DUI
ARREST TIME: 23:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:13

SUBJECT'S LAST NAME: MASON

DATE/TIME OF TEST: 09/05/2017 01:08:33

Diagnostics Check	OK	01:01
Air Blank	0.000	01:01
Control Test	0.077	01:01
Air Blank	0.000	01:02
Subject Sample # 1	0.111	01:03
Breath Volume	1.355	01:03
Air Blank	0.000	01:03
Air Blank	0.000	01:05
Subject Sample # 2	0.117	01:07
Breath Volume	2.468	01:07
Air Blank	0.000	01:07
Control Test	0.080	01:07
Air Blank	0.000	01:08
Diagnostics Check	OK	01:08

Data Download Date/Time: 09/28/2017 17:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001292

INSTRUMENT REGISTERED TO: LEVY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINKOFER,ZACHARY,R
BREATH TEST OPERATOR AGENCY: Levy County SO
ARRESTING OFFICER: ADAMS,M
ARRESTING OFFICER AGENCY: Levy County SO
LAST AGENCY INSPECTION: 08/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG619605
DRY GAS EXPIRY DATE: 07/14/2018
VIOLATION CODE: DUI
ARREST TIME: 21:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: WILES

DATE/TIME OF TEST: 09/11/2017 22:00:56

Diagnostics Check	OK	21:58	
Air Blank	0.000	21:58	
Control Test	0.079	21:58	
Air Blank	0.000	21:59	
Subject Sample	# 1	21:59	Subject Test Refused
Breath Volume	0.000	21:59	
Air Blank	0.000	22:00	
Control Test	0.079	22:00	
Air Blank	0.000	22:00	
Diagnostics Check	OK	22:00	

Data Download Date/Time: 09/28/2017 17:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001292

INSTRUMENT REGISTERED TO: LEVY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COSSEBOOM,M,J
BREATH TEST OPERATOR AGENCY: Chiefland PD
ARRESTING OFFICER: COSSEBOOM,M
ARRESTING OFFICER AGENCY: Chiefland PD
LAST AGENCY INSPECTION: 08/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG619605
DRY GAS EXPIRY DATE: 07/14/2018
VIOLATION CODE: DUI
ARREST TIME: 21:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: BURNS

DATE/TIME OF TEST: 09/17/2017 22:36:39

Diagnostics Check	OK	22:33	
Air Blank	0.000	22:33	
Control Test	0.079	22:33	
Air Blank	0.000	22:34	
Subject Sample	# 1	22:35	Subject Test Refused
Breath Volume	0.000	22:35	
Air Blank	0.000	22:35	
Control Test	0.079	22:36	
Air Blank	0.000	22:36	
Diagnostics Check	OK	22:36	

Data Download Date/Time: 09/28/2017 17:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001292

INSTRUMENT REGISTERED TO: LEVY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COSSEBOOM,M,J
BREATH TEST OPERATOR AGENCY: Chiefland PD
ARRESTING OFFICER: COSSEBOOM,M
ARRESTING OFFICER AGENCY: Chiefland PD
LAST AGENCY INSPECTION: 08/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG619605
DRY GAS EXPIRY DATE: 07/14/2018
VIOLATION CODE: DUI
ARREST TIME: 01:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:47

SUBJECT'S LAST NAME: HULON JR

DATE/TIME OF TEST: 09/22/2017 03:12:42

Diagnostics Check	OK	03:08	
Air Blank	0.000	03:08	
Control Test	0.079	03:08	
Air Blank	0.000	03:09	
Subject Sample	# 1	03:11	Subject Test Refused
Breath Volume	0.000	03:11	
Air Blank	0.000	03:11	
Control Test	0.079	03:12	
Air Blank	0.000	03:12	
Diagnostics Check	OK	03:12	

Data Download Date/Time: 09/28/2017 17:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001292

INSTRUMENT REGISTERED TO: LEVY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LESHNER, TIFFANY, A
BREATH TEST OPERATOR AGENCY: Levy County SO
ARRESTING OFFICER: MASHBURN, J
ARRESTING OFFICER AGENCY: Levy County SO
LAST AGENCY INSPECTION: 08/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG619605
DRY GAS EXPIRY DATE: 07/14/2018
VIOLATION CODE: DUI
ARREST TIME: 00:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: KNIGHT

DATE/TIME OF TEST: 09/24/2017 02:12:43

Diagnostics Check	OK	02:06
Air Blank	0.000	02:06
Control Test	0.079	02:07
Air Blank	0.000	02:07
Subject Sample # 1	0.131	02:08
Breath Volume	2.195	02:08
Air Blank	0.000	02:08
Air Blank	0.000	02:10
Subject Sample # 2	0.128	02:11
Breath Volume	1.726	02:11
Air Blank	0.000	02:11
Control Test	0.079	02:12
Air Blank	0.000	02:12
Diagnostics Check	OK	02:12

Data Download Date/Time: 09/28/2017 17:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001293

INSTRUMENT REGISTERED TO: WILLISTON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANDIS,BRYAN,S
BREATH TEST OPERATOR AGENCY: Williston PD
ARRESTING OFFICER: LANDIS,BRYAN
ARRESTING OFFICER AGENCY: Williston PD
LAST AGENCY INSPECTION: 08/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 13:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:46

SUBJECT'S LAST NAME: BERENGER

DATE/TIME OF TEST: 09/03/2017 14:37:20

Diagnostics Check	OK	14:31
Air Blank	0.000	14:31
Control Test	0.079	14:31
Air Blank	0.000	14:32
Subject Sample # 1	0.000	14:32
Breath Volume	2.867	14:32
Air Blank	0.000	14:33
Air Blank	0.000	14:35
Subject Sample # 2	0.000	14:35
Breath Volume	2.257	14:35
Air Blank	0.000	14:36
Control Test	0.078	14:36
Air Blank	0.000	14:37
Diagnostics Check	OK	14:37

Data Download Date/Time: 09/23/2017 11:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001293

INSTRUMENT REGISTERED TO: WILLISTON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANDIS,BRYAN,S
BREATH TEST OPERATOR AGENCY: Williston PD
ARRESTING OFFICER: PEOPLES,NATHAN
ARRESTING OFFICER AGENCY: Williston PD
LAST AGENCY INSPECTION: 09/23/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 20:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 09/29/2017 21:38:46

Diagnostics Check	OK	21:31
Air Blank	0.000	21:32
Control Test	0.079	21:32
Air Blank	0.000	21:33
Subject Sample # 1	0.199	21:33
Breath Volume	1.566	21:33
Air Blank	0.000	21:34
Air Blank	0.000	21:36
Subject Sample # 2	0.209	21:37
Breath Volume	2.226	21:37
Air Blank	0.000	21:37
Control Test	0.079	21:38
Air Blank	0.000	21:38
Diagnostics Check	OK	21:38

Data Download Date/Time: 12/13/2017 11:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001293

INSTRUMENT REGISTERED TO: WILLISTON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANDIS,BRYAN,L
BREATH TEST OPERATOR AGENCY: Williston PD
ARRESTING OFFICER: WRIGHT,MARQUIS
ARRESTING OFFICER AGENCY: Williston PD
LAST AGENCY INSPECTION: 09/23/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 03:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:13

SUBJECT'S LAST NAME: TORRES

DATE/TIME OF TEST: 09/30/2017 05:39:52

Diagnostics Check	OK	05:32
Air Blank	0.000	05:33
Control Test	0.078	05:33
Air Blank	0.000	05:34
Subject Sample # 1	0.101	05:34
Breath Volume	1.964	05:34
Air Blank	0.000	05:35
Air Blank	0.000	05:37
Subject Sample # 2	0.100	05:38
Breath Volume	2.359	05:38
Air Blank	0.000	05:38
Control Test	0.078	05:39
Air Blank	0.000	05:39
Diagnostics Check	OK	05:39

Data Download Date/Time: 12/13/2017 11:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACKSON,VICTOR,A
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: KHUNE,ROBERT
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 09/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629801
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: OTHER
ARREST TIME: 11:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:30

SUBJECT'S LAST NAME: ANDREU

DATE/TIME OF TEST: 09/28/2017 14:09:04

Diagnostics Check	OK	14:02
Air Blank	0.000	14:03
Control Test	0.079	14:03
Air Blank	0.000	14:04
Subject Sample # 1	0.000	14:04
Breath Volume	2.292	14:04
Air Blank	0.000	14:05
Air Blank	0.000	14:07
Subject Sample # 2	0.000	14:07
Breath Volume	2.117	14:07
Air Blank	0.000	14:08
Control Test	0.080	14:08
Air Blank	0.000	14:08
Diagnostics Check	OK	14:09

Data Download Date/Time: 10/31/2017 18:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHAPMAN,MAVERICK,E
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: CHAPMAN,MAVERICK
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 09/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629801
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 21:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: BARE

DATE/TIME OF TEST: 09/29/2017 22:21:40

Diagnostics Check	OK	22:10	
Air Blank	0.000	22:10	
Control Test	0.080	22:11	
Air Blank	0.000	22:11	
Subject Sample	# 1	22:14	No Sample Provided
Breath Volume	0.000	22:14	
Air Blank	0.000	22:15	
Air Blank	0.000	22:17	
Subject Sample	# 2	22:20	No Sample Provided
Breath Volume	0.000	22:20	
Air Blank	0.000	22:20	
Control Test	0.080	22:21	
Air Blank	0.000	22:21	
Diagnostics Check	OK	22:21	

Data Download Date/Time: 10/31/2017 18:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CALABRO,ALFRED,J
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: CALABRO,ALFRED
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 09/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629801
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 19:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:15

SUBJECT'S LAST NAME: KING

DATE/TIME OF TEST: 09/30/2017 20:42:35

Diagnostics Check	OK	20:36
Air Blank	0.000	20:36
Control Test	0.078	20:36
Air Blank	0.000	20:37
Subject Sample # 1	0.206	20:38
Breath Volume	1.726	20:38
Air Blank	0.000	20:38
Air Blank	0.000	20:40
Subject Sample # 2	0.201	20:41
Breath Volume	1.691	20:41
Air Blank	0.000	20:41
Control Test	0.078	20:41
Air Blank	0.000	20:42
Diagnostics Check	OK	20:42

Data Download Date/Time: 10/31/2017 18:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARVOTH,DANIEL,P
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: HARVOTH,DANIEL
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 09/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629801
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 21:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:01

SUBJECT'S LAST NAME: ELLIS

DATE/TIME OF TEST: 09/30/2017 22:30:22

Diagnostics Check	OK	22:21	
Air Blank	0.000	22:22	
Control Test	0.079	22:22	
Air Blank	0.000	22:23	
Subject Sample	# 1	22:26	No Sample Provided
Breath Volume	0.000	22:26	
Air Blank	0.000	22:26	
Air Blank	0.000	22:28	
Subject Sample	# 2	22:29	Subject Test Refused
Breath Volume	0.000	22:29	
Air Blank	0.000	22:29	
Control Test	0.079	22:29	
Air Blank	0.000	22:30	
Diagnostics Check	OK	22:30	

Data Download Date/Time: 10/31/2017 18:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MIDYETTE,COREY,L
BREATH TEST OPERATOR AGENCY: Parker PD
ARRESTING OFFICER: JOHNSON,JUSTIN
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 08/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG628401
DRY GAS EXPIRY DATE: 10/10/2018
VIOLATION CODE: DUI
ARREST TIME: 23:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:57

SUBJECT'S LAST NAME: SIMPSON

DATE/TIME OF TEST: 09/03/2017 01:28:06

Diagnostics Check	OK	01:22
Air Blank	0.000	01:22
Control Test	0.079	01:22
Air Blank	0.000	01:23
Subject Sample # 1	0.147	01:23
Breath Volume	1.531	01:23
Air Blank	0.000	01:24
Air Blank	0.000	01:26
Subject Sample # 2	0.160	01:26
Breath Volume	1.921	01:26
Air Blank	0.000	01:27
Control Test	0.080	01:27
Air Blank	0.000	01:28
Diagnostics Check	OK	01:28

Data Download Date/Time: 09/25/2017 11:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ABRAHAM,ERIC,D
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: ABRAHAM,ERIC
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 08/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG628401
DRY GAS EXPIRY DATE: 10/10/2018
VIOLATION CODE: DUI
ARREST TIME: 12:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:56

SUBJECT'S LAST NAME: BAILEY JR

DATE/TIME OF TEST: 09/09/2017 13:49:34

Diagnostics Check	OK	13:36
Air Blank	0.000	13:36
Control Test	0.079	13:37
Air Blank	0.000	13:37
Subject Sample # 1	0.187	13:38
Breath Volume	1.746	13:38
Air Blank	0.000	13:39
Air Blank	0.000	13:41
Subject Sample # 2	0.158	13:43
Breath Volume	1.195	13:43
Air Blank	0.000	13:44
Air Blank	0.000	13:46
Subject Sample # 3	0.171	13:48
Breath Volume	1.882	13:48
Air Blank	0.000	13:48
Control Test	0.079	13:48
Air Blank	0.000	13:49
Diagnostics Check	OK	13:49

Data Download Date/Time: 09/25/2017 11:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ABRAHAM,ERIC,D
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: BERNAU,JOSHEPH
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 08/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG628401
DRY GAS EXPIRY DATE: 10/10/2018
VIOLATION CODE: DUI
ARREST TIME: 14:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:19

SUBJECT'S LAST NAME: ALLEN

DATE/TIME OF TEST: 09/09/2017 15:05:35

Diagnostics Check	OK	14:58
Air Blank	0.000	14:59
Control Test	0.080	14:59
Air Blank	0.000	15:00
Subject Sample # 1	0.000	15:00
Breath Volume	1.308	15:00
Air Blank	0.000	15:01
Air Blank	0.000	15:03
Subject Sample # 2	0.000	15:04
Breath Volume	1.546	15:04
Air Blank	0.000	15:04
Control Test	0.079	15:05
Air Blank	0.000	15:05
Diagnostics Check	OK	15:05

Data Download Date/Time: 09/25/2017 11:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHIPMAN,KAREN,D
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: CHASTAIN,KRISTA
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 08/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG628401
DRY GAS EXPIRY DATE: 10/10/2018
VIOLATION CODE: DUI
ARREST TIME: 01:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: VILLEGAS

DATE/TIME OF TEST: 09/10/2017 03:00:47

Diagnostics Check	OK	02:51	
Air Blank	0.000	02:51	
Control Test	0.079	02:52	
Air Blank	0.000	02:52	
Subject Sample	# 1 0.117	02:55	Volume Not Met
Breath Volume	0.984	02:55	
Air Blank	0.000	02:56	
Air Blank	0.000	02:58	
Subject Sample	# 2	02:59	Subject Test Refused
Breath Volume	0.617	02:59	
Air Blank	0.000	02:59	
Control Test	0.078	03:00	
Air Blank	0.000	03:00	
Diagnostics Check	OK	03:00	

Data Download Date/Time: 09/25/2017 11:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOOD,ERIC,R
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: WOOD,ERIC
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 08/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG628401
DRY GAS EXPIRY DATE: 10/10/2018
VIOLATION CODE: DUI
ARREST TIME: 18:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:25

SUBJECT'S LAST NAME: CAPPETTA

DATE/TIME OF TEST: 09/18/2017 20:08:13

Diagnostics Check	OK	20:00
Air Blank	0.000	20:01
Control Test	0.079	20:01
Air Blank	0.000	20:02
Subject Sample # 1	0.251	20:02
Breath Volume	1.148	20:02
Air Blank	0.000	20:03
Air Blank	0.000	20:05
Subject Sample # 2	0.239	20:06
Breath Volume	2.246	20:06
Air Blank	0.000	20:07
Control Test	0.076	20:07
Air Blank	0.000	20:08
Diagnostics Check	OK	20:08

Data Download Date/Time: 09/25/2017 11:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOBER,JOSHUA,P
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: PARKER,CHRISTY
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 08/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG628401
DRY GAS EXPIRY DATE: 10/10/2018
VIOLATION CODE: DUI
ARREST TIME: 01:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:22

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 09/23/2017 01:49:58

Diagnostics Check	OK	01:45	
Air Blank	0.000	01:46	
Control Test	0.079	01:46	
Air Blank	0.000	01:47	
Subject Sample	# 1	01:48	Subject Test Refused
Breath Volume	0.000	01:48	
Air Blank	0.000	01:49	
Control Test	0.079	01:49	
Air Blank	0.000	01:49	
Diagnostics Check	OK	01:49	

Data Download Date/Time: 09/25/2017 11:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRICE,CARLON,E
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: GOBER,JOSHUA
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 08/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG628401
DRY GAS EXPIRY DATE: 10/10/2018
VIOLATION CODE: DUI
ARREST TIME: 00:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:48

SUBJECT'S LAST NAME: MORIARTY

DATE/TIME OF TEST: 09/24/2017 02:11:00

Diagnostics Check	OK	02:08	
Air Blank	0.000	02:08	
Control Test	0.079	02:08	
Air Blank	0.000	02:09	
Subject Sample	# 1	02:09	Subject Test Refused
Breath Volume	0.000	02:09	
Air Blank	0.000	02:10	
Control Test	0.079	02:10	
Air Blank	0.000	02:10	
Diagnostics Check	OK	02:11	

Data Download Date/Time: 09/25/2017 11:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,STEPHANIE,N
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: MATHIS,MARKEVIS
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 08/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG628401
DRY GAS EXPIRY DATE: 10/10/2018
VIOLATION CODE: DUI
ARREST TIME: 01:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: CHISOLM

DATE/TIME OF TEST: 09/02/2017 02:26:27

Diagnostics Check		OK	02:20
Air Blank		0.000	02:20
Control Test		0.077	02:21
Air Blank		0.000	02:21
Subject Sample	# 1	0.116	02:22
Breath Volume		1.714	02:22
Air Blank		0.000	02:22
Air Blank		0.000	02:24
Subject Sample	# 2	0.115	02:24
Breath Volume		1.765	02:24
Air Blank		0.000	02:25
Control Test		0.078	02:25
Air Blank		0.000	02:26
Diagnostics Check		OK	02:26

Data Download Date/Time: 09/25/2017 11:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOTZONG,DUSTIN,R
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: BOTZONG,DUSTIN
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 08/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG628401
DRY GAS EXPIRY DATE: 10/10/2018
VIOLATION CODE: DUI
ARREST TIME: 17:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:33

SUBJECT'S LAST NAME: FOSTER

DATE/TIME OF TEST: 09/08/2017 18:56:28

Diagnostics Check	OK	18:53	
Air Blank	0.000	18:53	
Control Test	0.077	18:54	
Air Blank	0.000	18:54	
Subject Sample	# 1	18:55	Subject Test Refused
Breath Volume	0.000	18:55	
Air Blank	0.000	18:55	
Control Test	0.077	18:55	
Air Blank	0.000	18:56	
Diagnostics Check	OK	18:56	

Data Download Date/Time: 09/25/2017 11:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOOD,ERIC,R
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: WOOD,ERIC
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 08/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG628401
DRY GAS EXPIRY DATE: 10/10/2018
VIOLATION CODE: DUI
ARREST TIME: 19:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:25

SUBJECT'S LAST NAME: CAPPETTA

DATE/TIME OF TEST: 09/18/2017 19:51:21

Diagnostics Check	OK	19:50	
Air Blank	0.000	19:50	
Control Test	0.078	19:50	
Air Blank		19:50	RFI Detect
Air Blank		19:51	RFI Detect

Data Download Date/Time: 09/25/2017 11:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRICE,CARLON,E
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: PRICE,CARLON
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 08/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG628401
DRY GAS EXPIRY DATE: 10/10/2018
VIOLATION CODE: DUI
ARREST TIME: 00:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:53

SUBJECT'S LAST NAME: GLASPELL

DATE/TIME OF TEST: 09/24/2017 02:26:37

Diagnostics Check	OK	02:20
Air Blank	0.000	02:20
Control Test	0.077	02:21
Air Blank	0.000	02:21
Subject Sample # 1	0.206	02:22
Breath Volume	2.019	02:22
Air Blank	0.000	02:22
Air Blank	0.000	02:24
Subject Sample # 2	0.220	02:25
Breath Volume	3.398	02:25
Air Blank	0.000	02:25
Control Test	0.076	02:26
Air Blank	0.000	02:26
Diagnostics Check	OK	02:26

Data Download Date/Time: 09/25/2017 11:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THORE,BECKY,A
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: BROCK,AUSTIN
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 09/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG628401
DRY GAS EXPIRY DATE: 10/10/2018
VIOLATION CODE: DUI
ARREST TIME: 03:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:03

SUBJECT'S LAST NAME: SHEPHERD

DATE/TIME OF TEST: 09/30/2017 04:24:09

Diagnostics Check	Fail	04:23	DSP Test
Air Blank	0.000	04:24	

Data Download Date/Time: 10/30/2017 10:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THORE,BECKY,A
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: BROCK,AUSTIN
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 09/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG628401
DRY GAS EXPIRY DATE: 10/10/2018
VIOLATION CODE: DUI
ARREST TIME: 03:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:03

SUBJECT'S LAST NAME: SHEPHERD

DATE/TIME OF TEST: 09/30/2017 04:38:37

Diagnostics Check	Fail	04:38	DSP Test
Air Blank	0.000	04:38	

Data Download Date/Time: 10/30/2017 10:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001310

INSTRUMENT REGISTERED TO: NICEVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MACBETH, MATTHEW, R
BREATH TEST OPERATOR AGENCY: Niceville PD
ARRESTING OFFICER: SULLIVAN, DANNY
ARRESTING OFFICER AGENCY: Niceville PD
LAST AGENCY INSPECTION: 09/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: DUI
ARREST TIME: 19:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: BARNES

DATE/TIME OF TEST: 09/30/2017 20:27:55

Diagnostics Check	OK	20:21
Air Blank	0.000	20:21
Control Test	0.082	20:22
Air Blank	0.000	20:22
Subject Sample # 1	0.162	20:23
Breath Volume	2.054	20:23
Air Blank	0.000	20:23
Air Blank	0.000	20:25
Subject Sample # 2	0.154	20:26
Breath Volume	1.621	20:26
Air Blank	0.000	20:27
Control Test	0.082	20:27
Air Blank	0.000	20:27
Diagnostics Check	OK	20:27

Data Download Date/Time: 12/03/2017 13:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AIMANOVICH, MATTHEW,
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: WHITEHEAD, JEFF
ARRESTING OFFICER AGENCY: DeFuniak Springs PD
LAST AGENCY INSPECTION: 08/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163937
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 20:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:29

SUBJECT'S LAST NAME: BROCK

DATE/TIME OF TEST: 09/01/2017 22:00:25

Diagnostics Check	OK	21:52
Air Blank	0.000	21:53
Control Test	0.080	21:53
Air Blank	0.000	21:54
Subject Sample # 1	0.225	21:55
Breath Volume	1.562	21:55
Air Blank	0.000	21:56
Air Blank	0.000	21:58
Subject Sample # 2	0.223	21:58
Breath Volume	1.707	21:58
Air Blank	0.000	21:59
Control Test	0.077	21:59
Air Blank	0.000	22:00
Diagnostics Check	OK	22:00

Data Download Date/Time: 09/30/2017 10:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUNTER,ROBERT,A
BREATH TEST OPERATOR AGENCY: DeFuniak Springs PD
ARRESTING OFFICER: HUNTER,ROBERT
ARRESTING OFFICER AGENCY: DeFuniak Springs PD
LAST AGENCY INSPECTION: 08/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 06:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:22

SUBJECT'S LAST NAME: TENORIO

DATE/TIME OF TEST: 09/16/2017 08:00:06

Diagnostics Check	OK	07:52
Air Blank	0.000	07:53
Control Test	0.081	07:53
Air Blank	0.000	07:54
Subject Sample # 1	0.190	07:54
Breath Volume	2.042	07:54
Air Blank	0.000	07:55
Air Blank	0.000	07:57
Subject Sample # 2	0.184	07:58
Breath Volume	3.335	07:58
Air Blank	0.000	07:59
Control Test	0.080	07:59
Air Blank	0.000	08:00
Diagnostics Check	OK	08:00

Data Download Date/Time: 09/30/2017 10:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LINCOLN,TREY,B
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: LINCOLN,TREY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 08/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163937
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 16:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:35

SUBJECT'S LAST NAME: GRANDE-MARTINEZ

DATE/TIME OF TEST: 09/24/2017 19:32:01

Diagnostics Check		OK	19:16	
Air Blank		0.000	19:16	
Control Test		0.081	19:16	
Air Blank		0.000	19:17	
Subject Sample	# 1	0.105	19:20	Volume Not Met
Breath Volume		0.683	19:20	
Air Blank		0.000	19:21	
Air Blank		0.000	19:23	
Subject Sample	# 2	0.136	19:24	
Breath Volume		1.800	19:24	
Air Blank		0.000	19:25	
Air Blank		0.000	19:27	
Subject Sample	# 3		19:30	No Sample Provided
Breath Volume		0.000	19:30	
Air Blank		0.000	19:31	
Control Test		0.078	19:31	
Air Blank		0.000	19:31	
Diagnostics Check		OK	19:32	

Data Download Date/Time: 09/30/2017 10:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FUGATE, MITCHELL, W
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: JOINER, J
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 08/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 00:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: ESTRADA VERGARA

DATE/TIME OF TEST: 09/30/2017 01:27:35

Diagnostics Check	OK	01:16	
Air Blank	0.000	01:16	
Control Test	0.079	01:16	
Air Blank	0.000	01:17	
Subject Sample	# 1	01:20	No Sample Provided
Breath Volume	0.000	01:20	
Air Blank	0.000	01:21	
Air Blank	0.000	01:23	
Subject Sample	# 2	01:26	No Sample Provided
Breath Volume	0.000	01:26	
Air Blank	0.000	01:26	
Control Test	0.081	01:26	
Air Blank	0.000	01:27	
Diagnostics Check	OK	01:27	

Data Download Date/Time: 09/30/2017 10:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHULTZ,JEREMY,J
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: SCHULTZ,JEREMY
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 08/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: DUI
ARREST TIME: 04:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:25

SUBJECT'S LAST NAME: GRIFFIN

DATE/TIME OF TEST: 09/09/2017 05:02:06

Diagnostics Check	OK	04:50	
Air Blank	0.000	04:51	
Control Test	0.078	04:51	
Air Blank	0.000	04:51	
Subject Sample	# 1	04:55	No Sample Provided
Breath Volume	0.000	04:55	
Air Blank	0.000	04:55	
Air Blank	0.000	04:57	
Subject Sample	# 2	05:00	No Sample Provided
Breath Volume	0.000	05:00	
Air Blank	0.000	05:01	
Control Test	0.080	05:01	
Air Blank	0.000	05:02	
Diagnostics Check	OK	05:02	

Data Download Date/Time: 09/30/2017 10:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHULTZ,JEREMY,J
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: SCHULTZ,JEREMY
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 08/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: DUI
ARREST TIME: 22:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: CLE?

DATE/TIME OF TEST: 09/09/2017 23:02:04

Diagnostics Check	OK	22:55
Air Blank	0.000	22:55
Control Test	0.077	22:55
Air Blank	0.000	22:56
Subject Sample # 1	0.175	22:57
Breath Volume	1.386	22:57
Air Blank	0.000	22:57
Air Blank	0.000	22:59
Subject Sample # 2	0.185	23:00
Breath Volume	2.773	23:00
Air Blank	0.000	23:01
Control Test	0.078	23:01
Air Blank	0.000	23:01
Diagnostics Check	OK	23:02

Data Download Date/Time: 09/30/2017 10:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GAYDOS,JONATHAN,A
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: GAYDOS,JONATHAN
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 09/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: PROB/PAROL
ARREST TIME: 13:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:10

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 09/30/2017 13:41:49

Diagnostics Check	OK	13:35	
Air Blank	0.000	13:36	
Control Test	0.079	13:36	
Air Blank	0.000	13:37	
Subject Sample # 1	0.000	13:37	
Breath Volume	1.945	13:37	
Air Blank	0.000	13:38	
Air Blank	0.000	13:40	
Subject Sample # 2	0.000	13:40	
Breath Volume	1.656	13:40	
Air Blank	0.000	13:40	
Control Test	0.082	13:41	Interferent Detect
Air Blank	0.000	13:41	

Data Download Date/Time: 10/10/2017 13:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001320

INSTRUMENT REGISTERED TO: OKEECHOBEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POLLOCK,STEVEN,D
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: POLLOCK,STEVEN
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 09/07/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:13

SUBJECT'S LAST NAME: FORD

DATE/TIME OF TEST: 09/14/2017 00:39:35

Diagnostics Check	OK	00:33
Air Blank	0.000	00:33
Control Test	0.080	00:34
Air Blank	0.000	00:34
Subject Sample # 1	0.226	00:35
Breath Volume	1.371	00:35
Air Blank	0.000	00:35
Air Blank	0.000	00:37
Subject Sample # 2	0.224	00:38
Breath Volume	2.289	00:38
Air Blank	0.000	00:38
Control Test	0.080	00:39
Air Blank	0.000	00:39
Diagnostics Check	OK	00:39

Data Download Date/Time: 10/04/2017 09:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001320

INSTRUMENT REGISTERED TO: OKEECHOBEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POLLOCK,STEVEN,D
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: POLLOCK,STEVEN
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 09/07/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: DUI
ARREST TIME: 00:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:29

SUBJECT'S LAST NAME: GARCIA GUZMAN

DATE/TIME OF TEST: 09/24/2017 00:53:17

Diagnostics Check	OK	00:50	
Air Blank	0.000	00:50	
Control Test	0.080	00:51	
Air Blank	0.000	00:51	
Subject Sample	# 1	00:51	Subject Test Refused
Breath Volume	0.000	00:51	
Air Blank	0.000	00:52	
Control Test	0.080	00:52	
Air Blank	0.000	00:53	
Diagnostics Check	OK	00:53	

Data Download Date/Time: 10/04/2017 09:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001320

INSTRUMENT REGISTERED TO: OKEECHOBEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POLLOCK,STEVEN,D
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: POLLOCK,STEVEN
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 09/07/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: DUI
ARREST TIME: 21:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:33

SUBJECT'S LAST NAME: MAYNOR

DATE/TIME OF TEST: 09/24/2017 22:01:48

Diagnostics Check	OK	21:53
Air Blank	0.000	21:54
Control Test	0.081	21:54
Air Blank	0.000	21:55
Subject Sample # 1	0.115	21:57
Breath Volume	1.914	21:57
Air Blank	0.000	21:57
Air Blank	0.000	21:59
Subject Sample # 2	0.120	22:00
Breath Volume	1.378	22:00
Air Blank	0.000	22:00
Control Test	0.080	22:01
Air Blank	0.000	22:01
Diagnostics Check	OK	22:01

Data Download Date/Time: 10/04/2017 09:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001322

INSTRUMENT REGISTERED TO: OKEECHOBEE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARRERO,RAUL,E
BREATH TEST OPERATOR AGENCY: Okeechobee PD
ARRESTING OFFICER: MARRERO,RAUL
ARRESTING OFFICER AGENCY: Okeechobee PD
LAST AGENCY INSPECTION: 08/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04216080A3
DRY GAS EXPIRY DATE: 03/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:32

SUBJECT'S LAST NAME: ELKINS

DATE/TIME OF TEST: 09/14/2017 23:04:18

Diagnostics Check	OK	22:57
Air Blank	0.000	22:57
Control Test	0.077	22:58
Air Blank	0.000	22:58
Subject Sample # 1	0.197	22:59
Breath Volume	2.585	22:59
Air Blank	0.000	23:00
Air Blank	0.000	23:02
Subject Sample # 2	0.197	23:02
Breath Volume	2.496	23:02
Air Blank	0.000	23:03
Control Test	0.076	23:03
Air Blank	0.000	23:04
Diagnostics Check	OK	23:04

Data Download Date/Time: 09/29/2017 10:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TREVISOL,JOSEPH,D
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: ZAMOYSKI,ADRIAN
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 08/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: YARBROUGH

DATE/TIME OF TEST: 09/11/2017 23:10:43

Diagnostics Check	OK	23:04
Air Blank	0.000	23:04
Control Test	0.079	23:04
Air Blank	0.000	23:05
Subject Sample # 1	0.280	23:06
Breath Volume	3.378	23:06
Air Blank	0.000	23:06
Air Blank	0.000	23:08
Subject Sample # 2	0.272	23:09
Breath Volume	2.890	23:09
Air Blank	0.000	23:09
Control Test	0.079	23:10
Air Blank	0.000	23:10
Diagnostics Check	OK	23:10

Data Download Date/Time: 09/28/2017 09:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERRING,TIMOTHY,R
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: HERRING,TIMOTHY
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 08/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: MENDEZ

DATE/TIME OF TEST: 09/14/2017 03:54:31

Diagnostics Check		OK	03:47
Air Blank		0.000	03:48
Control Test		0.080	03:48
Air Blank		0.000	03:48
Subject Sample	# 1	0.202	03:50
Breath Volume		1.992	03:50
Air Blank		0.000	03:50
Air Blank		0.000	03:52
Subject Sample	# 2	0.197	03:52
Breath Volume		1.789	03:52
Air Blank		0.000	03:53
Control Test		0.078	03:53
Air Blank		0.000	03:54
Diagnostics Check		OK	03:54

Data Download Date/Time: 09/28/2017 09:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TREVISOL,JOSEPH,D
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: TREVISOL,JOSEPH
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 08/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 19:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:35

SUBJECT'S LAST NAME: LE ANDRE

DATE/TIME OF TEST: 09/14/2017 20:04:21

Diagnostics Check	OK	20:01	
Air Blank	0.000	20:01	
Control Test	0.079	20:02	
Air Blank	0.000	20:02	
Subject Sample	# 1	20:03	Subject Test Refused
Breath Volume	0.000	20:03	
Air Blank	0.000	20:03	
Control Test	0.079	20:03	
Air Blank	0.000	20:04	
Diagnostics Check	OK	20:04	

Data Download Date/Time: 09/28/2017 09:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUEDE,KAYLA,M
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: HARDISON,AISHA
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 08/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:28

SUBJECT'S LAST NAME: FERGUSON

DATE/TIME OF TEST: 09/16/2017 01:52:25

Diagnostics Check	OK	01:49	
Air Blank	0.000	01:50	
Control Test	0.080	01:50	
Air Blank	0.000	01:50	
Subject Sample	# 1	01:51	Subject Test Refused
Breath Volume	0.000	01:51	
Air Blank	0.000	01:51	
Control Test	0.081	01:51	
Air Blank	0.000	01:52	
Diagnostics Check	OK	01:52	

Data Download Date/Time: 09/28/2017 09:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERRING,TIMOTHY,R
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: HERRING,TIMOTHY
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 08/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: MOORE

DATE/TIME OF TEST: 09/21/2017 03:40:04

Diagnostics Check	OK	03:26	
Air Blank	0.000	03:26	
Control Test	0.081	03:26	
Air Blank	0.000	03:27	
Subject Sample	# 1 0.163	03:29	
Breath Volume	1.292	03:29	
Air Blank	0.000	03:30	
Air Blank	0.000	03:31	
Subject Sample	# 2 0.145	03:35	Volume Not Met
Breath Volume	0.847	03:35	
Air Blank	0.000	03:35	
Air Blank	0.000	03:37	
Subject Sample	# 3 0.164	03:38	
Breath Volume	1.617	03:38	
Air Blank	0.000	03:39	
Control Test	0.076	03:39	
Air Blank	0.000	03:39	
Diagnostics Check	OK	03:40	

Data Download Date/Time: 09/28/2017 09:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,SALVADOR,
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: HARDISON,AISHA
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 08/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 19:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:51

SUBJECT'S LAST NAME: D520471551640

DATE/TIME OF TEST: 09/22/2017 20:14:02

Diagnostics Check	OK	20:12	
Air Blank	0.000	20:12	
Control Test	0.080	20:13	
Air Blank		20:13	RFI Detect
Air Blank	0.000	20:14	

Data Download Date/Time: 09/28/2017 09:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,SALVADOR,
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: HARDISON,AISHA
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 08/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 19:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:51

SUBJECT'S LAST NAME: DENYKO

DATE/TIME OF TEST: 09/22/2017 20:23:48

Diagnostics Check	OK	20:15	
Air Blank	0.000	20:15	
Control Test	0.080	20:16	
Air Blank	0.000	20:16	
Subject Sample	# 1	20:19	No Sample Provided
Breath Volume	0.000	20:19	
Air Blank	0.000	20:20	
Air Blank	0.000	20:22	
Subject Sample	# 2	20:22	Subject Test Refused
Breath Volume	0.000	20:22	
Air Blank	0.000	20:22	
Control Test	0.080	20:23	
Air Blank	0.000	20:23	
Diagnostics Check	OK	20:23	

Data Download Date/Time: 09/28/2017 09:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,SALVADOR,
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: PENNEY,ROBERT
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 08/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 19:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:33

SUBJECT'S LAST NAME: LE ANDRE

DATE/TIME OF TEST: 09/23/2017 20:00:47

Diagnostics Check	OK	19:53
Air Blank	0.000	19:54
Control Test	0.080	19:54
Air Blank	0.000	19:55
Subject Sample # 1	0.137	19:56
Breath Volume	1.792	19:56
Air Blank	0.000	19:56
Air Blank	0.000	19:58
Subject Sample # 2	0.139	19:59
Breath Volume	1.386	19:59
Air Blank	0.000	19:59
Control Test	0.078	20:00
Air Blank	0.000	20:00
Diagnostics Check	OK	20:00

Data Download Date/Time: 09/28/2017 09:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,SALVADOR,
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: DENNIS,ARIEL
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 08/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:17

SUBJECT'S LAST NAME: ANGELONE

DATE/TIME OF TEST: 09/23/2017 22:49:50

Diagnostics Check	OK	22:43
Air Blank	0.000	22:44
Control Test	0.080	22:44
Air Blank	0.000	22:45
Subject Sample # 1	0.225	22:45
Breath Volume	2.914	22:45
Air Blank	0.000	22:46
Air Blank	0.000	22:47
Subject Sample # 2	0.231	22:48
Breath Volume	3.503	22:48
Air Blank	0.000	22:48
Control Test	0.079	22:49
Air Blank	0.000	22:49
Diagnostics Check	OK	22:49

Data Download Date/Time: 09/28/2017 09:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUEDE,KAYLA,M
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: RUEDE,KAYLA
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 08/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: COLTRELL

DATE/TIME OF TEST: 09/24/2017 01:12:22

Diagnostics Check	OK	01:09	
Air Blank	0.000	01:10	
Control Test	0.080	01:10	
Air Blank	0.000	01:10	
Subject Sample	# 1	01:11	Subject Test Refused
Breath Volume	0.000	01:11	
Air Blank	0.000	01:11	
Control Test	0.080	01:11	
Air Blank	0.000	01:12	
Diagnostics Check	OK	01:12	

Data Download Date/Time: 09/28/2017 09:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUEDE,KAYLA,M
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: MCINERNEY,JESSE
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 08/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:39

SUBJECT'S LAST NAME: LOMAS

DATE/TIME OF TEST: 09/26/2017 04:03:33

Diagnostics Check	OK	04:00	
Air Blank	0.000	04:01	
Control Test	0.080	04:01	
Air Blank	0.000	04:02	
Subject Sample	# 1	04:02	Subject Test Refused
Breath Volume	0.000	04:02	
Air Blank	0.000	04:02	
Control Test	0.080	04:02	
Air Blank	0.000	04:03	
Diagnostics Check	OK	04:03	

Data Download Date/Time: 09/28/2017 09:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHIMENTI,RICHARD,J
BREATH TEST OPERATOR AGENCY: Vero Beach PD
ARRESTING OFFICER: EDER,KYLE
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 08/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 684131
DRY GAS EXPIRY DATE: 01/23/2020
VIOLATION CODE: DUI
ARREST TIME: 10:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 11:10

SUBJECT'S LAST NAME: TOMMIE

DATE/TIME OF TEST: 09/02/2017 11:55:30

Diagnostics Check		OK	11:41	
Air Blank		0.000	11:41	
Control Test		0.080	11:42	
Air Blank		0.000	11:42	
Subject Sample	# 1	0.170	11:45	Volume Not Met
Breath Volume		0.464	11:45	
Air Blank		0.000	11:46	
Air Blank		0.000	11:48	
Subject Sample	# 2	0.180	11:50	
Breath Volume		1.609	11:50	
Air Blank		0.000	11:51	
Air Blank		0.000	11:52	
Subject Sample	# 3	0.186	11:53	
Breath Volume		1.585	11:53	
Air Blank		0.000	11:54	
Control Test		0.078	11:54	
Air Blank		0.000	11:55	
Diagnostics Check		OK	11:55	

Data Download Date/Time: 09/25/2017 05:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ, MAEGAN,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: EVANS, ZACH
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 08/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 684131
DRY GAS EXPIRY DATE: 01/23/2020
VIOLATION CODE: DUI
ARREST TIME: 18:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:10

SUBJECT'S LAST NAME: SATDANOV

DATE/TIME OF TEST: 09/02/2017 20:06:28

Diagnostics Check	OK	19:57	
Air Blank	0.000	19:57	
Control Test	0.080	19:58	
Air Blank	0.000	19:58	
Subject Sample	# 1 0.000	20:01	Volume Not Met
Breath Volume	0.023	20:01	
Air Blank	0.000	20:02	
Air Blank	0.000	20:04	
Subject Sample	# 2	20:05	Subject Test Refused
Breath Volume	0.000	20:05	
Air Blank	0.000	20:05	
Control Test	0.079	20:05	
Air Blank	0.000	20:06	
Diagnostics Check	OK	20:06	

Data Download Date/Time: 09/25/2017 05:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ, MAEGAN,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: EVANS, ZACHARY
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 08/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 684131
DRY GAS EXPIRY DATE: 01/23/2020
VIOLATION CODE: DUI
ARREST TIME: 18:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:10

SUBJECT'S LAST NAME: SATDANOV

DATE/TIME OF TEST: 09/02/2017 20:12:10

Diagnostics Check	OK	20:09	
Air Blank	0.000	20:09	
Control Test	0.081	20:09	
Air Blank	0.000	20:10	
Subject Sample	# 1	20:10	Subject Test Refused
Breath Volume	0.000	20:10	
Air Blank	0.000	20:11	
Control Test	0.080	20:11	
Air Blank	0.000	20:12	
Diagnostics Check	OK	20:12	

Data Download Date/Time: 09/25/2017 05:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GITTO,JUSTIN,J
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: GITTO,JUSTIN
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 08/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 684131
DRY GAS EXPIRY DATE: 01/23/2020
VIOLATION CODE: DUI
ARREST TIME: 22:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:37

SUBJECT'S LAST NAME: MOTINO

DATE/TIME OF TEST: 09/03/2017 23:04:47

Diagnostics Check	OK	22:57
Air Blank	0.000	22:58
Control Test	0.080	22:58
Air Blank	0.000	22:58
Subject Sample # 1	0.211	23:00
Breath Volume	1.554	23:00
Air Blank	0.000	23:01
Air Blank	0.000	23:02
Subject Sample # 2	0.212	23:03
Breath Volume	1.539	23:03
Air Blank	0.000	23:03
Control Test	0.079	23:04
Air Blank	0.000	23:04
Diagnostics Check	OK	23:04

Data Download Date/Time: 09/25/2017 05:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,LEE,A
BREATH TEST OPERATOR AGENCY: Vero Beach PD
ARRESTING OFFICER: EVANS,LEE
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 08/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 684131
DRY GAS EXPIRY DATE: 01/23/2020
VIOLATION CODE: DUI
ARREST TIME: 01:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:44

SUBJECT'S LAST NAME: KOBY

DATE/TIME OF TEST: 09/04/2017 02:11:49

Diagnostics Check	OK	02:05
Air Blank	0.000	02:05
Control Test	0.081	02:06
Air Blank	0.000	02:06
Subject Sample # 1	0.131	02:07
Breath Volume	4.535	02:07
Air Blank	0.000	02:08
Air Blank	0.000	02:09
Subject Sample # 2	0.121	02:10
Breath Volume	4.136	02:10
Air Blank	0.000	02:10
Control Test	0.079	02:11
Air Blank	0.000	02:11
Diagnostics Check	OK	02:11

Data Download Date/Time: 09/25/2017 05:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIST,SHARON,D
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: PARKER,CARLIN
ARRESTING OFFICER AGENCY: Sebastian PD
LAST AGENCY INSPECTION: 08/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 684131
DRY GAS EXPIRY DATE: 01/23/2020
VIOLATION CODE: DUI
ARREST TIME: 11:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 12:50

SUBJECT'S LAST NAME: SNOW

DATE/TIME OF TEST: 09/06/2017 13:18:04

Diagnostics Check	OK	13:12	
Air Blank	0.000	13:13	
Control Test	0.079	13:13	
Air Blank	0.000	13:14	
Subject Sample	# 1	13:16	Subject Test Refused
Breath Volume	0.484	13:16	
Air Blank	0.000	13:17	
Control Test	0.081	13:17	
Air Blank	0.000	13:17	
Diagnostics Check	OK	13:18	

Data Download Date/Time: 09/25/2017 05:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIST,SHARON,D
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: PARKER,CALIN
ARRESTING OFFICER AGENCY: Sebastian PD
LAST AGENCY INSPECTION: 08/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 684131
DRY GAS EXPIRY DATE: 01/23/2020
VIOLATION CODE: DUI
ARREST TIME: 11:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 12:50

SUBJECT'S LAST NAME: SNOW

DATE/TIME OF TEST: 09/06/2017 13:29:09

Diagnostics Check	OK	13:21
Air Blank	0.000	13:21
Control Test	0.080	13:21
Air Blank	0.000	13:22
Subject Sample # 1	0.160	13:23
Breath Volume	1.789	13:23
Air Blank	0.000	13:24
Air Blank	0.000	13:26
Subject Sample # 2	0.149	13:27
Breath Volume	1.117	13:27
Air Blank	0.000	13:28
Control Test	0.078	13:28
Air Blank	0.000	13:29
Diagnostics Check	OK	13:29

Data Download Date/Time: 09/25/2017 05:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ, MAEGAN,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: EGGERS, R
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 08/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 684131
DRY GAS EXPIRY DATE: 01/23/2020
VIOLATION CODE: DUI
ARREST TIME: 17:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:13

SUBJECT'S LAST NAME: KASPER

DATE/TIME OF TEST: 09/07/2017 18:46:26

Diagnostics Check	OK	18:39
Air Blank	0.000	18:40
Control Test	0.080	18:40
Air Blank	0.000	18:40
Subject Sample # 1	0.253	18:41
Breath Volume	1.144	18:41
Air Blank	0.000	18:42
Air Blank	0.000	18:44
Subject Sample # 2	0.269	18:44
Breath Volume	1.234	18:44
Air Blank	0.000	18:45
Control Test	0.079	18:45
Air Blank	0.000	18:46
Diagnostics Check	OK	18:46

Data Download Date/Time: 09/25/2017 05:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STOVALL,C,J
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: THOMAS,L
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 08/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 684131
DRY GAS EXPIRY DATE: 01/23/2020
VIOLATION CODE: DUI
ARREST TIME: 19:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:40

SUBJECT'S LAST NAME: LEWICKI

DATE/TIME OF TEST: 09/07/2017 20:35:17

Diagnostics Check	OK	20:32	
Air Blank	0.000	20:32	
Control Test	0.081	20:33	
Air Blank	0.000	20:33	
Subject Sample	# 1	20:33	Subject Test Refused
Breath Volume	0.000	20:33	
Air Blank	0.000	20:34	
Control Test	0.080	20:34	
Air Blank	0.000	20:35	
Diagnostics Check	OK	20:35	

Data Download Date/Time: 09/25/2017 05:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUZMAN,MELISSA,M
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: PIANELLI,VICTORIA
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 08/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 684131
DRY GAS EXPIRY DATE: 01/23/2020
VIOLATION CODE: DUI
ARREST TIME: 16:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:56

SUBJECT'S LAST NAME: ARNOLD

DATE/TIME OF TEST: 09/08/2017 18:25:04

Diagnostics Check	OK	18:18
Air Blank	0.000	18:19
Control Test	0.081	18:19
Air Blank	0.000	18:19
Subject Sample # 1	0.325	18:20
Breath Volume	1.917	18:20
Air Blank	0.000	18:21
Air Blank	0.000	18:23
Subject Sample # 2	0.339	18:23
Breath Volume	1.285	18:23
Air Blank	0.000	18:24
Control Test	0.079	18:24
Air Blank	0.000	18:24
Diagnostics Check	OK	18:25

Data Download Date/Time: 09/25/2017 05:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALLAGHER, MATTHEW, T
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: GALLAGHER, MATTHEW
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 08/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 684131
DRY GAS EXPIRY DATE: 01/23/2020
VIOLATION CODE: DUI
ARREST TIME: 18:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:42

SUBJECT'S LAST NAME: DUQUESNE

DATE/TIME OF TEST: 09/08/2017 19:05:18

Diagnostics Check	OK	19:02	
Air Blank	0.000	19:02	
Control Test	0.080	19:03	
Air Blank	0.000	19:03	
Subject Sample	# 1	19:03	Subject Test Refused
Breath Volume	0.000	19:03	
Air Blank	0.000	19:04	
Control Test	0.080	19:04	
Air Blank	0.000	19:05	
Diagnostics Check	OK	19:05	

Data Download Date/Time: 09/25/2017 05:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOLSKI,JEFFREY,R
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: COREY,T
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 08/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 684131
DRY GAS EXPIRY DATE: 01/23/2020
VIOLATION CODE: DUI
ARREST TIME: 15:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:35

SUBJECT'S LAST NAME: PETERSON

DATE/TIME OF TEST: 09/09/2017 17:08:08

Diagnostics Check	OK	16:56	
Air Blank	0.000	16:57	
Control Test	0.080	16:57	
Air Blank	0.000	16:58	
Subject Sample	# 1	17:01	No Sample Provided
Breath Volume	0.000	17:01	
Air Blank	0.000	17:01	
Air Blank	0.000	17:03	
Subject Sample	# 2	17:06	No Sample Provided
Breath Volume	0.000	17:06	
Air Blank	0.000	17:07	
Control Test	0.080	17:07	
Air Blank	0.000	17:08	
Diagnostics Check	OK	17:08	

Data Download Date/Time: 09/25/2017 05:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOLSKI,JEFFREY,R
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: COREY,T
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 08/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 684131
DRY GAS EXPIRY DATE: 01/23/2020
VIOLATION CODE: DUI
ARREST TIME: 15:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:35

SUBJECT'S LAST NAME: PETERSON

DATE/TIME OF TEST: 09/09/2017 17:13:48

Diagnostics Check	OK	17:10	
Air Blank	0.000	17:11	
Control Test	0.080	17:11	
Air Blank	0.000	17:12	
Subject Sample	# 1	17:12	Subject Test Refused
Breath Volume	0.000	17:12	
Air Blank	0.000	17:12	
Control Test	0.080	17:13	
Air Blank	0.000	17:13	
Diagnostics Check	OK	17:13	

Data Download Date/Time: 09/25/2017 05:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SWINDELL,CHRISTOPHER,E
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: HATFIELD,CHAYSE
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 08/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 684131
DRY GAS EXPIRY DATE: 01/23/2020
VIOLATION CODE: DUI
ARREST TIME: 00:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:19

SUBJECT'S LAST NAME: SNODGRESS

DATE/TIME OF TEST: 09/12/2017 01:09:45

Diagnostics Check		OK	00:57	
Air Blank		0.000	00:57	
Control Test		0.080	00:57	
Air Blank		0.000	00:58	
Subject Sample	# 1	0.142	01:01	Volume Not Met
Breath Volume		0.878	01:01	
Air Blank		0.000	01:02	
Air Blank		0.000	01:04	
Subject Sample	# 2	0.177	01:05	
Breath Volume		1.882	01:05	
Air Blank		0.000	01:05	
Air Blank		0.000	01:07	
Subject Sample	# 3	0.177	01:08	
Breath Volume		1.660	01:08	
Air Blank		0.000	01:08	
Control Test		0.079	01:09	
Air Blank		0.000	01:09	
Diagnostics Check		OK	01:09	

Data Download Date/Time: 09/25/2017 05:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,LEE,A
BREATH TEST OPERATOR AGENCY: Vero Beach PD
ARRESTING OFFICER: EVANS,LEE
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 08/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 684131
DRY GAS EXPIRY DATE: 01/23/2020
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:23

SUBJECT'S LAST NAME: BELL

DATE/TIME OF TEST: 09/12/2017 03:13:39

Diagnostics Check	OK	03:10	
Air Blank	0.000	03:10	
Control Test	0.079	03:11	
Air Blank	0.000	03:11	
Subject Sample	# 1	03:12	Subject Test Refused
Breath Volume	0.000	03:12	
Air Blank	0.000	03:12	
Control Test	0.080	03:13	
Air Blank	0.000	03:13	
Diagnostics Check	OK	03:13	

Data Download Date/Time: 09/25/2017 05:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMILTON,DEVIN,C
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: SHARPETA,CHAD
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 08/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 684131
DRY GAS EXPIRY DATE: 01/23/2020
VIOLATION CODE: DUI
ARREST TIME: 00:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:47

SUBJECT'S LAST NAME: OWENS

DATE/TIME OF TEST: 09/15/2017 02:43:38

Diagnostics Check	OK	02:34
Air Blank	0.000	02:34
Control Test	0.080	02:34
Air Blank	0.000	02:35
Subject Sample # 1	0.122	02:36
Breath Volume	1.226	02:36
Air Blank	0.000	02:37
Air Blank	0.000	02:39
Subject Sample # 2	0.128	02:42
Breath Volume	1.871	02:42
Air Blank	0.000	02:42
Control Test	0.081	02:43
Air Blank	0.000	02:43
Diagnostics Check	OK	02:43

Data Download Date/Time: 09/25/2017 05:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YORK, JIMMY, A
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: FRICKE, JOANNA
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 08/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 684131
DRY GAS EXPIRY DATE: 01/23/2020
VIOLATION CODE: DUI
ARREST TIME: 02:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:58

SUBJECT'S LAST NAME: STRICKLAND

DATE/TIME OF TEST: 09/17/2017 03:22:48

Diagnostics Check	OK	03:19	
Air Blank	0.000	03:20	
Control Test	0.080	03:20	
Air Blank	0.000	03:21	
Subject Sample	# 1	03:21	Subject Test Refused
Breath Volume	0.000	03:21	
Air Blank	0.000	03:21	
Control Test	0.080	03:22	
Air Blank	0.000	03:22	
Diagnostics Check	OK	03:22	

Data Download Date/Time: 09/25/2017 05:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STOVALL,C,J
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: SINGH,T
ARRESTING OFFICER AGENCY: Sebastian PD
LAST AGENCY INSPECTION: 08/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 684131
DRY GAS EXPIRY DATE: 01/23/2020
VIOLATION CODE: DUI
ARREST TIME: 00:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: JETT

DATE/TIME OF TEST: 09/18/2017 01:41:04

Diagnostics Check	OK	01:38	
Air Blank	0.000	01:38	
Control Test	0.080	01:39	
Air Blank	0.000	01:39	
Subject Sample	# 1	01:39	Subject Test Refused
Breath Volume	0.000	01:39	
Air Blank	0.000	01:40	
Control Test	0.080	01:40	
Air Blank	0.000	01:40	
Diagnostics Check	OK	01:41	

Data Download Date/Time: 09/25/2017 05:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,QUINTON,T
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: EGGERS,RYAN
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 08/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 684131
DRY GAS EXPIRY DATE: 01/23/2020
VIOLATION CODE: DUI
ARREST TIME: 19:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:50

SUBJECT'S LAST NAME: LEPINE

DATE/TIME OF TEST: 09/19/2017 20:27:39

Diagnostics Check	OK	20:21
Air Blank	0.000	20:21
Control Test	0.080	20:22
Air Blank	0.000	20:22
Subject Sample # 1	0.000	20:23
Breath Volume	1.191	20:23
Air Blank	0.000	20:23
Air Blank	0.000	20:25
Subject Sample # 2	0.000	20:26
Breath Volume	1.386	20:26
Air Blank	0.000	20:26
Control Test	0.081	20:27
Air Blank	0.000	20:27
Diagnostics Check	OK	20:27

Data Download Date/Time: 09/25/2017 05:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STOVALL,C,J
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: DELGADO,J
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 08/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 684131
DRY GAS EXPIRY DATE: 01/23/2020
VIOLATION CODE: DUI
ARREST TIME: 21:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: O BRIEN

DATE/TIME OF TEST: 09/23/2017 22:46:25

Diagnostics Check	OK	22:43	
Air Blank	0.000	22:44	
Control Test	0.081	22:44	
Air Blank	0.000	22:44	
Subject Sample	# 1	22:45	Subject Test Refused
Breath Volume	0.000	22:45	
Air Blank	0.000	22:45	
Control Test	0.079	22:45	
Air Blank	0.000	22:46	
Diagnostics Check	OK	22:46	

Data Download Date/Time: 09/25/2017 05:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOLSKI,JEFFREY,R
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: REDSTONE,M
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 09/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 684131
DRY GAS EXPIRY DATE: 01/23/2020
VIOLATION CODE: DUI
ARREST TIME: 15:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:27

SUBJECT'S LAST NAME: MARTIN

DATE/TIME OF TEST: 09/25/2017 16:54:01

Diagnostics Check	OK	16:51	
Air Blank	0.000	16:51	
Control Test	0.080	16:51	
Air Blank	0.000	16:52	
Subject Sample	# 1	16:52	Subject Test Refused
Breath Volume	0.000	16:52	
Air Blank	0.000	16:53	
Control Test	0.080	16:53	
Air Blank	0.000	16:53	
Diagnostics Check	OK	16:54	

Data Download Date/Time: 10/13/2017 05:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,QUINTON,T
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MATTON,NICKOLAS
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 09/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 684131
DRY GAS EXPIRY DATE: 01/23/2020
VIOLATION CODE: DUI
ARREST TIME: 21:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:25

SUBJECT'S LAST NAME: PERKINS

DATE/TIME OF TEST: 09/26/2017 23:00:57

Diagnostics Check	OK	22:51	
Air Blank	0.000	22:51	
Control Test	0.080	22:51	
Air Blank	0.000	22:52	
Subject Sample	# 1	22:55	No Sample Provided
Breath Volume	0.000	22:55	
Air Blank	0.000	22:56	
Air Blank	0.000	22:58	
Subject Sample	# 2	22:59	Subject Test Refused
Breath Volume	0.000	22:59	
Air Blank	0.000	23:00	
Control Test	0.080	23:00	
Air Blank	0.000	23:00	
Diagnostics Check	OK	23:00	

Data Download Date/Time: 10/13/2017 05:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOOD, THERESA, K
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MUNNIA
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 09/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 684131
DRY GAS EXPIRY DATE: 01/23/2020
VIOLATION CODE: DUI
ARREST TIME: 12:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:08

SUBJECT'S LAST NAME: SPEAKMAN

DATE/TIME OF TEST: 09/27/2017 14:54:57

Diagnostics Check	OK	14:52	
Air Blank	0.000	14:52	
Control Test	0.080	14:52	
Air Blank	0.000	14:53	
Subject Sample	# 1	14:53	RFI Detect
Breath Volume	0.000	14:53	
Air Blank	0.000	14:54	
Control Test	0.080	14:54	
Air Blank	0.000	14:54	
Diagnostics Check	OK	14:54	

Data Download Date/Time: 10/13/2017 05:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOOD, THERESA, K
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MUNNIA, A
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 09/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 684131
DRY GAS EXPIRY DATE: 01/23/2020
VIOLATION CODE: DUI
ARREST TIME: 12:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:08

SUBJECT'S LAST NAME: SPEAKMAN

DATE/TIME OF TEST: 09/27/2017 15:03:03

Diagnostics Check	OK	14:56
Air Blank	0.000	14:56
Control Test	0.079	14:57
Air Blank	0.000	14:57
Subject Sample # 1	0.000	14:58
Breath Volume	1.414	14:58
Air Blank	0.000	14:58
Air Blank	0.000	15:00
Subject Sample # 2	0.000	15:01
Breath Volume	1.648	15:01
Air Blank	0.000	15:02
Control Test	0.080	15:02
Air Blank	0.000	15:02
Diagnostics Check	OK	15:03

Data Download Date/Time: 10/13/2017 05:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMILTON,DEVIN,C
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: BAKER,MARCUS
ARRESTING OFFICER AGENCY: Indian River Shores
LAST AGENCY INSPECTION: 09/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 684131
DRY GAS EXPIRY DATE: 01/23/2020
VIOLATION CODE: DUI
ARREST TIME: 23:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:24

SUBJECT'S LAST NAME: MORALES

DATE/TIME OF TEST: 09/29/2017 01:57:34

Diagnostics Check	OK	01:54	
Air Blank	0.000	01:55	
Control Test	0.079	01:55	
Air Blank	0.000	01:55	
Subject Sample	# 1	01:56	Subject Test Refused
Breath Volume	0.000	01:56	
Air Blank	0.000	01:56	
Control Test	0.080	01:56	
Air Blank	0.000	01:57	
Diagnostics Check	OK	01:57	

Data Download Date/Time: 10/13/2017 05:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOLSKI,JEFFREY,R
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: NIEVES,L
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 09/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 684131
DRY GAS EXPIRY DATE: 01/23/2020
VIOLATION CODE: DUI
ARREST TIME: 18:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:40

SUBJECT'S LAST NAME: IERCZCHOWSKI\$ARTHUR\$

DATE/TIME OF TEST: 09/29/2017 19:09:03

Diagnostics Check		OK	19:02
Air Blank		0.000	19:03
Control Test		0.081	19:03
Air Blank		0.000	19:04
Subject Sample	# 1	0.269	19:04
Breath Volume		2.625	19:04
Air Blank		0.000	19:05
Air Blank		0.000	19:07
Subject Sample	# 2	0.264	19:07
Breath Volume		2.238	19:07
Air Blank		0.000	19:08
Control Test		0.079	19:08
Air Blank		0.000	19:08
Diagnostics Check		OK	19:09

Data Download Date/Time: 10/13/2017 05:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001341

INSTRUMENT REGISTERED TO: DESOTO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HORTON,EDMOND,W
BREATH TEST OPERATOR AGENCY: DeSoto County SO
ARRESTING OFFICER: LAMBERT,CASEY
ARRESTING OFFICER AGENCY: Arcadia PD
LAST AGENCY INSPECTION: 08/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:31

SUBJECT'S LAST NAME: C636445500590

DATE/TIME OF TEST: 09/07/2017 21:00:38

Diagnostics Check	OK	20:53
Air Blank	0.000	20:54
Control Test	0.079	20:54
Air Blank	0.000	20:55
Subject Sample # 1	0.190	20:55
Breath Volume	1.507	20:55
Air Blank	0.000	20:56
Air Blank	0.000	20:58
Subject Sample # 2	0.196	20:59
Breath Volume	1.816	20:59
Air Blank	0.000	20:59
Control Test	0.080	21:00
Air Blank	0.000	21:00
Diagnostics Check	OK	21:00

Data Download Date/Time: 09/22/2017 21:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001341

INSTRUMENT REGISTERED TO: DESOTO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HORTON,EDMOND,W
BREATH TEST OPERATOR AGENCY: DeSoto County SO
ARRESTING OFFICER: BAKER,DUSTIN
ARRESTING OFFICER AGENCY: DeSoto County SO
LAST AGENCY INSPECTION: 08/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:24

SUBJECT'S LAST NAME: SUTTON

DATE/TIME OF TEST: 09/15/2017 23:53:46

Diagnostics Check	OK	23:47
Air Blank	0.000	23:48
Control Test	0.079	23:48
Air Blank	0.000	23:48
Subject Sample # 1	0.136	23:49
Breath Volume	1.683	23:49
Air Blank	0.000	23:49
Air Blank	0.000	23:51
Subject Sample # 2	0.145	23:52
Breath Volume	1.507	23:52
Air Blank	0.000	23:52
Control Test	0.079	23:53
Air Blank	0.000	23:53
Diagnostics Check	OK	23:53

Data Download Date/Time: 09/22/2017 21:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001341

INSTRUMENT REGISTERED TO: DESOTO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PITTS,JOSHUA,E
BREATH TEST OPERATOR AGENCY: DeSoto County SO
ARRESTING OFFICER: PITTS,JOSHUA
ARRESTING OFFICER AGENCY: DeSoto County SO
LAST AGENCY INSPECTION: 09/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: PROB/PAROL
ARREST TIME: 12:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 12:04

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 09/25/2017 12:30:25

Diagnostics Check	OK	12:24
Air Blank	0.000	12:24
Control Test	0.079	12:25
Air Blank	0.000	12:25
Subject Sample # 1	0.000	12:26
Breath Volume	1.796	12:26
Air Blank	0.000	12:26
Air Blank	0.000	12:28
Subject Sample # 2	0.000	12:29
Breath Volume	1.902	12:29
Air Blank	0.000	12:29
Control Test	0.079	12:29
Air Blank	0.000	12:30
Diagnostics Check	OK	12:30

Data Download Date/Time: 10/20/2017 09:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUBRANO,DEP,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: SLOWICK,C
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 08/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 16:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:45

SUBJECT'S LAST NAME: CONNER

DATE/TIME OF TEST: 09/01/2017 18:23:03

Diagnostics Check	OK	18:12
Air Blank	0.000	18:13
Control Test	0.080	18:13
Air Blank	0.000	18:13
Subject Sample # 1	0.132	18:14
Breath Volume	2.089	18:14
Air Blank	0.000	18:15
Air Blank	0.000	18:17
Subject Sample # 2	0.109	18:17
Breath Volume	1.843	18:17
Air Blank	0.000	18:18
Air Blank	0.000	18:20
Subject Sample # 3	0.135	18:21
Breath Volume	4.082	18:21
Air Blank	0.000	18:22
Control Test	0.080	18:22
Air Blank	0.000	18:22
Diagnostics Check	OK	18:23

Data Download Date/Time: 09/27/2017 07:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARTHUR,MELVIN,J
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: ARTHUR,MELVIN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 08/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:06

SUBJECT'S LAST NAME: PRESLEY

DATE/TIME OF TEST: 09/02/2017 23:38:53

Diagnostics Check	OK	23:32
Air Blank	0.000	23:32
Control Test	0.080	23:33
Air Blank	0.000	23:33
Subject Sample # 1	0.161	23:34
Breath Volume	2.046	23:34
Air Blank	0.000	23:34
Air Blank	0.000	23:36
Subject Sample # 2	0.165	23:37
Breath Volume	1.804	23:37
Air Blank	0.000	23:37
Control Test	0.080	23:38
Air Blank	0.000	23:38
Diagnostics Check	OK	23:38

Data Download Date/Time: 09/27/2017 07:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SOBAS,PATRICIA,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: MARKEY,L
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 08/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:12

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 09/03/2017 01:44:03

Diagnostics Check	OK	01:37
Air Blank	0.000	01:37
Control Test	0.080	01:37
Air Blank	0.000	01:38
Subject Sample # 1	0.221	01:39
Breath Volume	1.785	01:39
Air Blank	0.000	01:39
Air Blank	0.000	01:41
Subject Sample # 2	0.224	01:42
Breath Volume	2.121	01:42
Air Blank	0.000	01:43
Control Test	0.079	01:43
Air Blank	0.000	01:43
Diagnostics Check	OK	01:44

Data Download Date/Time: 09/27/2017 07:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PUFAHL,G,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: PUFAHL,G
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 08/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:04

SUBJECT'S LAST NAME: HAYES

DATE/TIME OF TEST: 09/05/2017 22:53:33

Diagnostics Check	OK	22:46
Air Blank	0.000	22:46
Control Test	0.078	22:47
Air Blank	0.000	22:47
Subject Sample # 1	0.266	22:48
Breath Volume	2.203	22:48
Air Blank	0.000	22:49
Air Blank	0.000	22:50
Subject Sample # 2	0.258	22:51
Breath Volume	1.636	22:51
Air Blank	0.000	22:52
Control Test	0.079	22:52
Air Blank	0.000	22:53
Diagnostics Check	OK	22:53

Data Download Date/Time: 09/27/2017 07:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PUFAHL,G,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: PUFAHL,G
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 08/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:37

SUBJECT'S LAST NAME: MATTHEWS

DATE/TIME OF TEST: 09/07/2017 04:04:32

Diagnostics Check		OK	03:58
Air Blank		0.000	03:58
Control Test		0.081	03:58
Air Blank		0.000	03:59
Subject Sample	# 1	0.236	03:59
Breath Volume		1.914	03:59
Air Blank		0.000	04:00
Air Blank		0.000	04:02
Subject Sample	# 2	0.220	04:02
Breath Volume		1.593	04:02
Air Blank		0.000	04:03
Control Test		0.079	04:03
Air Blank		0.000	04:04
Diagnostics Check		OK	04:04

Data Download Date/Time: 09/27/2017 07:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KERR,CALEB,W
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: KERR,CALEB
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 08/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:21

SUBJECT'S LAST NAME: DEPTUS

DATE/TIME OF TEST: 09/12/2017 04:57:26

Diagnostics Check		OK	04:50
Air Blank		0.000	04:50
Control Test		0.080	04:50
Air Blank		0.000	04:51
Subject Sample	# 1	0.154	04:52
Breath Volume		2.507	04:52
Air Blank		0.000	04:53
Air Blank		0.000	04:55
Subject Sample	# 2	0.156	04:55
Breath Volume		2.097	04:55
Air Blank		0.000	04:56
Control Test		0.080	04:56
Air Blank		0.000	04:57
Diagnostics Check		OK	04:57

Data Download Date/Time: 09/27/2017 07:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARR,JEREMY,S
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: KIEFER,Z
ARRESTING OFFICER AGENCY: North Port PD
LAST AGENCY INSPECTION: 08/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 18:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:43

SUBJECT'S LAST NAME: SHEAD

DATE/TIME OF TEST: 09/12/2017 20:20:41

Diagnostics Check	OK	20:14
Air Blank	0.000	20:14
Control Test	0.079	20:15
Air Blank	0.000	20:15
Subject Sample # 1	0.254	20:16
Breath Volume	4.105	20:16
Air Blank	0.000	20:16
Air Blank	0.000	20:18
Subject Sample # 2	0.265	20:18
Breath Volume	4.175	20:18
Air Blank	0.000	20:19
Control Test	0.078	20:20
Air Blank	0.000	20:20
Diagnostics Check	OK	20:20

Data Download Date/Time: 09/27/2017 07:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSS, MICHAEL, K
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: LUNDSTROM, OFFICER
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 08/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 12:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 12:35

SUBJECT'S LAST NAME: TAPIA

DATE/TIME OF TEST: 09/14/2017 13:05:37

Diagnostics Check	OK	12:59
Air Blank	0.000	13:00
Control Test	0.080	13:00
Air Blank	0.000	13:00
Subject Sample # 1	0.000	13:01
Breath Volume	1.308	13:01
Air Blank	0.000	13:01
Air Blank	0.000	13:03
Subject Sample # 2	0.000	13:04
Breath Volume	1.863	13:04
Air Blank	0.000	13:04
Control Test	0.080	13:05
Air Blank	0.000	13:05
Diagnostics Check	OK	13:05

Data Download Date/Time: 09/27/2017 07:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FOWLER, TYLER, J
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: FOWLER, TYLER
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 08/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:24

SUBJECT'S LAST NAME: LAWHORN

DATE/TIME OF TEST: 09/15/2017 03:55:04

Diagnostics Check	OK	03:46
Air Blank	0.000	03:46
Control Test	0.080	03:47
Air Blank	0.000	03:47
Subject Sample # 1	0.167	03:50
Breath Volume	4.750	03:50
Air Blank	0.000	03:51
Air Blank	0.000	03:52
Subject Sample # 2	0.147	03:53
Breath Volume	1.281	03:53
Air Blank	0.000	03:54
Control Test	0.078	03:54
Air Blank	0.000	03:54
Diagnostics Check	OK	03:55

Data Download Date/Time: 09/27/2017 07:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,N,N
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: SLOWICK,C
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 08/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 12:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:54

SUBJECT'S LAST NAME: BUTLER

DATE/TIME OF TEST: 09/16/2017 02:32:38

Diagnostics Check		OK	02:22
Air Blank		0.000	02:22
Control Test		0.079	02:23
Air Blank		0.000	02:23
Subject Sample	# 1	0.209	02:25
Breath Volume		1.949	02:25
Air Blank		0.000	02:26
Air Blank		0.000	02:27
Subject Sample	# 2	0.200	02:30
Breath Volume		1.660	02:30
Air Blank		0.000	02:31
Control Test		0.076	02:31
Air Blank		0.000	02:32
Diagnostics Check		OK	02:32

Data Download Date/Time: 09/27/2017 07:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,DANIEL,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: SLOWICK,FHP
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 08/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 09/20/2017 23:55:01

Diagnostics Check	OK	23:46	
Air Blank	0.000	23:47	
Control Test	0.080	23:47	
Air Blank	0.000	23:47	
Subject Sample	# 1 0.127	23:51	Volume Not Met
Breath Volume	0.851	23:51	
Air Blank	0.000	23:51	
Air Blank	0.000	23:53	
Subject Sample	# 2	23:53	Subject Test Refused
Breath Volume	0.000	23:53	
Air Blank	0.000	23:54	
Control Test	0.079	23:54	
Air Blank	0.000	23:54	
Diagnostics Check	OK	23:55	

Data Download Date/Time: 09/27/2017 07:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,DANIEL,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: SLOWICK,FHP
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 08/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 09/21/2017 00:21:27

Diagnostics Check	OK	00:14
Air Blank	0.000	00:14
Control Test	0.081	00:15
Air Blank	0.000	00:15
Subject Sample # 1	0.139	00:16
Breath Volume	1.320	00:16
Air Blank	0.000	00:17
Air Blank	0.000	00:19
Subject Sample # 2	0.129	00:19
Breath Volume	1.257	00:19
Air Blank	0.000	00:20
Control Test	0.077	00:20
Air Blank	0.000	00:21
Diagnostics Check	OK	00:21

Data Download Date/Time: 09/27/2017 07:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURLEY,SHAUN,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: SMALL,K
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 08/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:57

SUBJECT'S LAST NAME: WELLS

DATE/TIME OF TEST: 09/23/2017 22:34:38

Diagnostics Check	OK	22:28
Air Blank	0.000	22:28
Control Test	0.080	22:28
Air Blank	0.000	22:29
Subject Sample # 1	0.249	22:29
Breath Volume	2.363	22:29
Air Blank	0.000	22:30
Air Blank	0.000	22:32
Subject Sample # 2	0.243	22:32
Breath Volume	2.265	22:32
Air Blank	0.000	22:33
Control Test	0.080	22:34
Air Blank	0.000	22:34
Diagnostics Check	OK	22:34

Data Download Date/Time: 09/27/2017 07:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUEHLER,DEP,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BUEHLER,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 08/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: HASSAB

DATE/TIME OF TEST: 09/24/2017 00:26:52

Diagnostics Check	OK	00:18
Air Blank	0.000	00:18
Control Test	0.080	00:18
Air Blank	0.000	00:19
Subject Sample # 1	0.098	00:21
Breath Volume	1.468	00:21
Air Blank	0.000	00:22
Air Blank	0.000	00:23
Subject Sample # 2	0.096	00:25
Breath Volume	2.027	00:25
Air Blank	0.000	00:25
Control Test	0.080	00:26
Air Blank	0.000	00:26
Diagnostics Check	OK	00:26

Data Download Date/Time: 09/27/2017 07:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARTHUR,MELVIN,J
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: ARTHUR,MELVIN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 08/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 16:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:41

SUBJECT'S LAST NAME: TRUESDELL

DATE/TIME OF TEST: 09/25/2017 18:14:21

Diagnostics Check	OK	18:02	
Air Blank	0.000	18:03	
Control Test	0.079	18:03	
Air Blank	0.000	18:03	
Subject Sample	# 1	18:07	No Sample Provided
Breath Volume	0.000	18:07	
Air Blank	0.000	18:07	
Air Blank	0.000	18:09	
Subject Sample	# 2	18:12	Volume Not Met
Breath Volume	0.082	18:12	
Air Blank	0.000	18:13	
Control Test	0.079	18:13	
Air Blank	0.000	18:14	
Diagnostics Check	OK	18:14	

Data Download Date/Time: 09/27/2017 07:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 08/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: BARBA

DATE/TIME OF TEST: 09/26/2017 03:16:51

Diagnostics Check	OK	03:07	
Air Blank	0.000	03:08	
Control Test	0.080	03:08	
Air Blank	0.000	03:09	
Subject Sample	# 1	03:12	No Sample Provided
Breath Volume	0.000	03:12	
Air Blank	0.000	03:12	
Air Blank	0.000	03:14	
Subject Sample	# 2	03:15	Subject Test Refused
Breath Volume	0.000	03:15	
Air Blank	0.000	03:15	
Control Test	0.081	03:16	
Air Blank	0.000	03:16	
Diagnostics Check	OK	03:16	

Data Download Date/Time: 09/27/2017 07:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TETREAU, RUSSELL, L
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: DEFRANCISCO, OFC
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 08/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 15:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:15

SUBJECT'S LAST NAME: TANKA

DATE/TIME OF TEST: 09/26/2017 16:59:37

Diagnostics Check	OK	16:53
Air Blank	0.000	16:53
Control Test	0.080	16:54
Air Blank	0.000	16:54
Subject Sample # 1	0.177	16:55
Breath Volume	1.550	16:55
Air Blank	0.000	16:55
Air Blank	0.000	16:57
Subject Sample # 2	0.193	16:57
Breath Volume	2.941	16:57
Air Blank	0.000	16:58
Control Test	0.079	16:59
Air Blank	0.000	16:59
Diagnostics Check	OK	16:59

Data Download Date/Time: 09/27/2017 07:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSTYAL,MATTHEW,A
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: KOSTYAL,MATTHEW
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 08/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: SWEEZEY

DATE/TIME OF TEST: 09/27/2017 03:14:22

Diagnostics Check	OK	03:07
Air Blank	0.000	03:08
Control Test	0.080	03:08
Air Blank	0.000	03:09
Subject Sample # 1	0.092	03:09
Breath Volume	3.648	03:09
Air Blank	0.000	03:10
Air Blank	0.000	03:12
Subject Sample # 2	0.091	03:12
Breath Volume	3.386	03:12
Air Blank	0.000	03:13
Control Test	0.080	03:13
Air Blank	0.000	03:14
Diagnostics Check	OK	03:14

Data Download Date/Time: 09/27/2017 07:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

INSTRUMENT REGISTERED TO: EUSTIS P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HACKETT,SEAN,J
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: BROOKS,JOSHUA
ARRESTING OFFICER AGENCY: Eustis PD
LAST AGENCY INSPECTION: 08/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 15:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:34

SUBJECT'S LAST NAME: TREMAIN

DATE/TIME OF TEST: 09/02/2017 15:59:02

Diagnostics Check	OK	15:58	
Air Blank	0.000	15:58	
		15:58	RFI Detect
Air Blank	0.000	15:59	

Data Download Date/Time: 09/05/2017 10:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

INSTRUMENT REGISTERED TO: EUSTIS P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HACKETT,SEAN,J
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: BROOKS,JOSHUA
ARRESTING OFFICER AGENCY: Eustis PD
LAST AGENCY INSPECTION: 08/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 15:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:34

SUBJECT'S LAST NAME: TREMAIN

DATE/TIME OF TEST: 09/02/2017 16:05:23

Diagnostics Check	OK	16:02	
Air Blank	0.000	16:03	
Control Test	0.078	16:03	
Air Blank	0.000	16:03	
Subject Sample	# 1	16:04	Subject Test Refused
Breath Volume	0.000	16:04	
Air Blank	0.000	16:04	
Control Test	0.077	16:04	
Air Blank	0.000	16:05	
Diagnostics Check	OK	16:05	

Data Download Date/Time: 09/05/2017 10:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

INSTRUMENT REGISTERED TO: EUSTIS P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWELL,SHAWNA,M
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: STARLING,WILLIAM
ARRESTING OFFICER AGENCY: Eustis PD
LAST AGENCY INSPECTION: 08/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 20:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:25

SUBJECT'S LAST NAME: HALL

DATE/TIME OF TEST: 09/04/2017 22:55:34

Diagnostics Check	OK	22:49
Air Blank	0.000	22:49
Control Test	0.079	22:50
Air Blank	0.000	22:50
Subject Sample # 1	0.170	22:51
Breath Volume	2.421	22:51
Air Blank	0.000	22:51
Air Blank	0.000	22:53
Subject Sample # 2	0.155	22:54
Breath Volume	2.820	22:54
Air Blank	0.000	22:54
Control Test	0.078	22:54
Air Blank	0.000	22:55
Diagnostics Check	OK	22:55

Data Download Date/Time: 09/05/2017 10:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

INSTRUMENT REGISTERED TO: EUSTIS P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWELL,SHAWNA,M
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: STARLING,WILLIAM
ARRESTING OFFICER AGENCY: Eustis PD
LAST AGENCY INSPECTION: 09/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 02:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: BRIDGEMAN

DATE/TIME OF TEST: 09/10/2017 03:34:07

Diagnostics Check	OK	03:27
Air Blank	0.000	03:28
Control Test	0.078	03:28
Air Blank	0.000	03:29
Subject Sample # 1	0.194	03:29
Breath Volume	2.046	03:29
Air Blank	0.000	03:30
Air Blank	0.000	03:32
Subject Sample # 2	0.201	03:32
Breath Volume	1.726	03:32
Air Blank	0.000	03:33
Control Test	0.078	03:33
Air Blank	0.000	03:34
Diagnostics Check	OK	03:34

Data Download Date/Time: 10/27/2017 14:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

INSTRUMENT REGISTERED TO: EUSTIS P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWELL,SHAWNA,M
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: HOWELL,SHAWNA
ARRESTING OFFICER AGENCY: Eustis PD
LAST AGENCY INSPECTION: 09/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 03:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: TODD

DATE/TIME OF TEST: 09/14/2017 04:37:25

Diagnostics Check	OK	04:31
Air Blank	0.000	04:31
Control Test	0.079	04:31
Air Blank	0.000	04:32
Subject Sample # 1	0.000	04:32
Breath Volume	1.101	04:32
Air Blank	0.000	04:33
Air Blank	0.000	04:35
Subject Sample # 2	0.000	04:36
Breath Volume	1.761	04:36
Air Blank	0.000	04:36
Control Test	0.080	04:36
Air Blank	0.000	04:37
Diagnostics Check	OK	04:37

Data Download Date/Time: 10/27/2017 14:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

INSTRUMENT REGISTERED TO: EUSTIS P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: STROUP,STEVE,A

BREATH TEST OPERATOR AGENCY: Eustis PD

ARRESTING OFFICER: KOUROUMOUSIS,GEORGE

ARRESTING OFFICER AGENCY: Eustis PD

LAST AGENCY INSPECTION: 09/05/2017

Date or Time Changed: Yes

DRY GAS STANDARD LOT #: 203498

DRY GAS EXPIRY DATE: 12/10/2018

VIOLATION CODE: DUI

ARREST TIME: 10:40

GENDER: Female

OBSERVATION PERIOD BEGAN: 11:11

SUBJECT'S LAST NAME: WHEATLEY

DATE/TIME OF TEST: 09/19/2017 11:42:13

Diagnostics Check	OK	11:35
Air Blank	0.000	11:36
Control Test	0.078	11:36
Air Blank	0.000	11:37
Subject Sample # 1	0.307	11:37
Breath Volume	2.687	11:37
Air Blank	0.000	11:38
Air Blank	0.000	11:40
Subject Sample # 2	0.293	11:40
Breath Volume	2.062	11:40
Air Blank	0.000	11:41
Control Test	0.075	11:41
Air Blank	0.000	11:42
Diagnostics Check	OK	11:42

Data Download Date/Time: 10/27/2017 14:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

INSTRUMENT REGISTERED TO: EUSTIS P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWELL,SHAWNA,M
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: CARPENTER,TRAVIS
ARRESTING OFFICER AGENCY: Eustis PD
LAST AGENCY INSPECTION: 09/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 00:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: BIRDSONG

DATE/TIME OF TEST: 09/23/2017 01:45:08

Diagnostics Check	OK	01:38
Air Blank	0.000	01:38
Control Test	0.078	01:39
Air Blank	0.000	01:39
Subject Sample # 1	0.057	01:40
Breath Volume	1.753	01:40
Air Blank	0.000	01:40
Air Blank	0.000	01:42
Subject Sample # 2	0.054	01:43
Breath Volume	2.210	01:43
Air Blank	0.000	01:44
Control Test	0.077	01:44
Air Blank	0.000	01:45
Diagnostics Check	OK	01:45

Data Download Date/Time: 10/27/2017 14:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

INSTRUMENT REGISTERED TO: EUSTIS P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILES,DANIELLE,N
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: LEVINGSTON,PHIL
ARRESTING OFFICER AGENCY: Eustis PD
LAST AGENCY INSPECTION: 09/05/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 18:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:22

SUBJECT'S LAST NAME: WAXMAN

DATE/TIME OF TEST: 09/29/2017 19:48:34

Diagnostics Check	OK	19:35	
Air Blank	0.000	19:36	
Control Test	0.078	19:36	
Air Blank	0.000	19:36	
Subject Sample	# 1 0.129	19:38	
Breath Volume	1.648	19:38	
Air Blank	0.000	19:39	
Air Blank	0.000	19:40	
Subject Sample	# 2 0.111	19:44	Volume Not Met
Breath Volume	0.109	19:44	
Air Blank	0.000	19:44	
Air Blank	0.000	19:46	
Subject Sample	# 3 0.122	19:46	
Breath Volume	1.597	19:46	
Air Blank	0.000	19:47	
Control Test	0.076	19:47	
Air Blank	0.000	19:48	
Diagnostics Check	OK	19:48	

Data Download Date/Time: 10/27/2017 14:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHUMANN,DEP,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: FLINT,SGT
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 08/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:08

SUBJECT'S LAST NAME: FUTCH

DATE/TIME OF TEST: 09/02/2017 22:41:22

Diagnostics Check	OK	22:35
Air Blank	0.000	22:35
Control Test	0.077	22:35
Air Blank	0.000	22:36
Subject Sample # 1	0.046	22:36
Breath Volume	4.210	22:36
Air Blank	0.000	22:37
Air Blank	0.000	22:39
Subject Sample # 2	0.043	22:39
Breath Volume	2.191	22:39
Air Blank	0.000	22:40
Control Test	0.077	22:40
Air Blank	0.000	22:41
Diagnostics Check	OK	22:41

Data Download Date/Time: 09/27/2017 07:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSTYAL,MATT,A
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: KOSTYAL,MATT
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 08/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 18:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:19

SUBJECT'S LAST NAME: DE MAMAN

DATE/TIME OF TEST: 09/03/2017 18:46:09

Diagnostics Check	OK	18:40
Air Blank	0.000	18:40
Control Test	0.078	18:40
Air Blank	0.000	18:41
Subject Sample # 1	0.214	18:41
Breath Volume	2.843	18:41
Air Blank	0.000	18:42
Air Blank	0.000	18:44
Subject Sample # 2	0.216	18:44
Breath Volume	2.320	18:44
Air Blank	0.000	18:45
Control Test	0.079	18:45
Air Blank	0.000	18:46
Diagnostics Check	OK	18:46

Data Download Date/Time: 09/27/2017 08:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSTYAL,MATT,A
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: KOSTYAL,MATT
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 08/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: NEAL

DATE/TIME OF TEST: 09/05/2017 02:26:32

Diagnostics Check	OK	02:20
Air Blank	0.000	02:20
Control Test	0.078	02:21
Air Blank	0.000	02:21
Subject Sample # 1	0.193	02:22
Breath Volume	1.910	02:22
Air Blank	0.000	02:22
Air Blank	0.000	02:24
Subject Sample # 2	0.196	02:25
Breath Volume	2.085	02:25
Air Blank	0.000	02:25
Control Test	0.077	02:25
Air Blank	0.000	02:26
Diagnostics Check	OK	02:26

Data Download Date/Time: 09/27/2017 08:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PUFAHL,G,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: KOSTYAL,M
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 08/30/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: SHERRY

DATE/TIME OF TEST: 09/07/2017 23:47:04

Diagnostics Check	OK	23:40
Air Blank	0.000	23:40
Control Test	0.078	23:40
Air Blank	0.000	23:41
Subject Sample # 1	0.165	23:42
Breath Volume	1.445	23:42
Air Blank	0.000	23:42
Air Blank	0.000	23:44
Subject Sample # 2	0.173	23:45
Breath Volume	1.988	23:45
Air Blank	0.000	23:46
Control Test	0.079	23:46
Air Blank	0.000	23:46
Diagnostics Check	OK	23:47

Data Download Date/Time: 09/27/2017 08:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSTYAL,MATT,A
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: KOSTYAL,MATT
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 08/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: SIMON

DATE/TIME OF TEST: 09/14/2017 00:47:11

Diagnostics Check		OK	00:41
Air Blank		0.000	00:41
Control Test		0.079	00:41
Air Blank		0.000	00:42
Subject Sample	# 1	0.281	00:42
Breath Volume		2.726	00:42
Air Blank		0.000	00:43
Air Blank		0.000	00:45
Subject Sample	# 2	0.284	00:45
Breath Volume		1.640	00:45
Air Blank		0.000	00:46
Control Test		0.079	00:46
Air Blank		0.000	00:47
Diagnostics Check		OK	00:47

Data Download Date/Time: 09/27/2017 08:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSTYAL,MATT,A
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: KOSTYAL,MATT
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 09/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: SALAZAR

DATE/TIME OF TEST: 09/29/2017 02:37:16

Diagnostics Check		OK	02:30
Air Blank		0.000	02:31
Control Test		0.078	02:31
Air Blank		0.000	02:32
Subject Sample	# 1	0.182	02:32
Breath Volume		2.691	02:32
Air Blank		0.000	02:33
Air Blank		0.000	02:34
Subject Sample	# 2	0.171	02:35
Breath Volume		1.328	02:35
Air Blank		0.000	02:36
Control Test		0.078	02:36
Air Blank		0.000	02:37
Diagnostics Check		OK	02:37

Data Download Date/Time: 10/25/2017 09:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCLAIN,CHRISTOPHER,M
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: LEUTHNER,N
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 08/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A6
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:33

SUBJECT'S LAST NAME: WAGNER

DATE/TIME OF TEST: 09/02/2017 02:02:43

Diagnostics Check	OK	01:53
Air Blank	0.000	01:54
Control Test	0.080	01:54
Air Blank	0.000	01:55
Subject Sample # 1	0.212	01:57
Breath Volume	2.328	01:57
Air Blank	0.000	01:58
Air Blank	0.000	02:00
Subject Sample # 2	0.214	02:01
Breath Volume	2.195	02:01
Air Blank	0.000	02:01
Control Test	0.079	02:02
Air Blank	0.000	02:02
Diagnostics Check	OK	02:02

Data Download Date/Time: 09/27/2017 11:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCLAIN,CHRISTOPHER,M
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: LEUTHNER,N
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 08/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A6
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: MCINTOSH

DATE/TIME OF TEST: 09/03/2017 01:31:08

Diagnostics Check	OK	01:23
Air Blank	0.000	01:23
Control Test	0.080	01:24
Air Blank	0.000	01:24
Subject Sample # 1	0.102	01:26
Breath Volume	2.257	01:26
Air Blank	0.000	01:26
Air Blank	0.000	01:28
Subject Sample # 2	0.105	01:29
Breath Volume	2.273	01:29
Air Blank	0.000	01:30
Control Test	0.080	01:30
Air Blank	0.000	01:31
Diagnostics Check	OK	01:31

Data Download Date/Time: 09/27/2017 11:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCLAIN,CHRISTOPHER,M
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: NIX,R
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 08/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A6
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 17:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:08

SUBJECT'S LAST NAME: DELOZIER

DATE/TIME OF TEST: 09/07/2017 18:59:44

Diagnostics Check		OK	18:46	
Air Blank		0.000	18:47	
Control Test		0.081	18:47	
Air Blank		0.000	18:47	
Subject Sample	# 1	0.226	18:50	Volume Not Met
Breath Volume		0.175	18:50	
Air Blank		0.000	18:51	
Air Blank		0.000	18:53	
Subject Sample	# 2	0.229	18:55	
Breath Volume		2.640	18:55	
Air Blank		0.000	18:55	
Air Blank		0.000	18:57	
Subject Sample	# 3	0.241	18:58	
Breath Volume		2.207	18:58	
Air Blank		0.000	18:58	
Control Test		0.078	18:59	
Air Blank		0.000	18:59	
Diagnostics Check		OK	18:59	

Data Download Date/Time: 09/27/2017 11:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHEPARD,C,A
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: COTE,J
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 08/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A6
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 13:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:02

SUBJECT'S LAST NAME: HUFF

DATE/TIME OF TEST: 09/09/2017 14:35:34

Diagnostics Check	OK	14:24	
Air Blank	0.000	14:24	
Control Test	0.080	14:24	
Air Blank	0.000	14:25	
Subject Sample	# 1 0.233	14:28	Volume Not Met
Breath Volume	0.613	14:28	
Air Blank	0.000	14:29	
Air Blank	0.000	14:30	
Subject Sample	# 2 0.209	14:34	Volume Not Met
Breath Volume	0.507	14:34	
Air Blank	0.000	14:34	
Control Test	0.071	14:34	Control Outside Tolerance
Air Blank	0.000	14:35	

Data Download Date/Time: 09/27/2017 11:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PLATZ,KEVIN,L
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: TRUFAN,JOE
ARRESTING OFFICER AGENCY: Punta Gorda PD
LAST AGENCY INSPECTION: 08/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A6
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: AMADOR GOMEZ

DATE/TIME OF TEST: 09/13/2017 00:13:07

Diagnostics Check	OK	00:05
Air Blank	0.000	00:06
Control Test	0.079	00:06
Air Blank	0.000	00:07
Subject Sample # 1	0.192	00:08
Breath Volume	2.386	00:08
Air Blank	0.000	00:08
Air Blank	0.000	00:10
Subject Sample # 2	0.196	00:11
Breath Volume	2.523	00:11
Air Blank	0.000	00:12
Control Test	0.080	00:12
Air Blank	0.000	00:13
Diagnostics Check	OK	00:13

Data Download Date/Time: 09/27/2017 11:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARKER, JEREMY, A
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: WRASSE, DEPUTY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 08/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A6
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:17

SUBJECT'S LAST NAME: LOVERIDGE

DATE/TIME OF TEST: 09/14/2017 23:45:48

Diagnostics Check	OK	23:39
Air Blank	0.000	23:40
Control Test	0.079	23:40
Air Blank	0.000	23:40
Subject Sample # 1	0.065	23:41
Breath Volume	1.574	23:41
Air Blank	0.000	23:42
Air Blank	0.000	23:44
Subject Sample # 2	0.060	23:44
Breath Volume	1.816	23:44
Air Blank	0.000	23:44
Control Test	0.080	23:45
Air Blank	0.000	23:45
Diagnostics Check	OK	23:45

Data Download Date/Time: 09/27/2017 11:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHAPMAN,VANESSA,N
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: CIGNARALE,P
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 08/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A6
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:01

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 09/16/2017 02:39:44

Diagnostics Check	OK	02:33
Air Blank	0.000	02:33
Control Test	0.080	02:34
Air Blank	0.000	02:34
Subject Sample # 1	0.000	02:35
Breath Volume	1.585	02:35
Air Blank	0.000	02:35
Air Blank	0.000	02:37
Subject Sample # 2	0.000	02:38
Breath Volume	2.187	02:38
Air Blank	0.000	02:38
Control Test	0.080	02:39
Air Blank	0.000	02:39
Diagnostics Check	OK	02:39

Data Download Date/Time: 09/27/2017 11:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EGAN,MELISSA,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: FREITAS,ANTHONY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 08/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A6
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: HANSON

DATE/TIME OF TEST: 09/17/2017 01:38:06

Diagnostics Check	OK	01:25	
Air Blank	0.000	01:26	
Control Test	0.080	01:26	
Air Blank	0.000	01:26	
Subject Sample	# 1 0.108	01:27	
Breath Volume	1.511	01:27	
Air Blank	0.000	01:28	
Air Blank	0.000	01:30	
Subject Sample	# 2 0.103	01:33	Volume Not Met
Breath Volume	0.347	01:33	
Air Blank	0.000	01:34	
Air Blank	0.000	01:36	
Subject Sample	# 3 0.104	01:36	
Breath Volume	1.812	01:36	
Air Blank	0.000	01:37	
Control Test	0.079	01:37	
Air Blank	0.000	01:38	
Diagnostics Check	OK	01:38	

Data Download Date/Time: 09/27/2017 11:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCLAIN,CHRISTOPHER,M
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: CORNIER,G
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 08/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A6
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 11:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:38

SUBJECT'S LAST NAME: SWAN

DATE/TIME OF TEST: 09/18/2017 01:16:28

Diagnostics Check		OK	01:04	
Air Blank		0.000	01:05	
Control Test		0.080	01:05	
Air Blank		0.000	01:05	
Subject Sample	# 1	0.167	01:09	Volume Not Met
Breath Volume		0.804	01:09	
Air Blank		0.000	01:09	
Air Blank		0.000	01:11	
Subject Sample	# 2	0.180	01:12	
Breath Volume		1.582	01:12	
Air Blank		0.000	01:12	
Air Blank		0.000	01:14	
Subject Sample	# 3	0.180	01:14	
Breath Volume		1.253	01:14	
Air Blank		0.000	01:15	
Control Test		0.079	01:15	
Air Blank		0.000	01:16	
Diagnostics Check		OK	01:16	

Data Download Date/Time: 09/27/2017 11:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEMEN,AARON,M
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: STEMEN,AARON
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 08/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:36

SUBJECT'S LAST NAME: GARDNER

DATE/TIME OF TEST: 09/01/2017 01:03:20

Diagnostics Check		OK	00:57
Air Blank		0.000	00:57
Control Test		0.079	00:57
Air Blank		0.000	00:58
Subject Sample	# 1	0.138	00:58
Breath Volume		3.656	00:58
Air Blank		0.000	00:59
Air Blank		0.000	01:01
Subject Sample	# 2	0.138	01:01
Breath Volume		3.414	01:01
Air Blank		0.000	01:02
Control Test		0.080	01:02
Air Blank		0.000	01:03
Diagnostics Check		OK	01:03

Data Download Date/Time: 09/21/2017 20:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRINSKO,THOMAS,K
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: GUYTON,JOHN
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 08/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:55

SUBJECT'S LAST NAME: DE JESUS LOZADA

DATE/TIME OF TEST: 09/02/2017 05:19:02

Diagnostics Check	OK	05:16	
Air Blank	0.000	05:16	
Control Test	0.079	05:16	
Air Blank	0.000	05:17	
Subject Sample	# 1	05:17	Subject Test Refused
Breath Volume	0.000	05:17	
Air Blank	0.000	05:18	
Control Test	0.080	05:18	
Air Blank	0.000	05:18	
Diagnostics Check	OK	05:19	

Data Download Date/Time: 09/21/2017 20:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEMEN,AARON,M
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: STEMEN,AARON
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 08/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:08

SUBJECT'S LAST NAME: BRYANT

DATE/TIME OF TEST: 09/02/2017 21:36:49

Diagnostics Check	OK	21:30
Air Blank	0.000	21:31
Control Test	0.079	21:31
Air Blank	0.000	21:31
Subject Sample # 1	0.258	21:32
Breath Volume	1.347	21:32
Air Blank	0.000	21:32
Air Blank	0.000	21:34
Subject Sample # 2	0.262	21:35
Breath Volume	1.273	21:35
Air Blank	0.000	21:35
Control Test	0.079	21:36
Air Blank	0.000	21:36
Diagnostics Check	OK	21:36

Data Download Date/Time: 09/21/2017 20:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEMEN,AARON,M
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: STEMEN,AARON
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 08/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: WADE

DATE/TIME OF TEST: 09/03/2017 03:42:31

Diagnostics Check	OK	03:33	
Air Blank	0.000	03:34	
Control Test	0.079	03:34	
Air Blank	0.000	03:35	
Subject Sample	# 1	03:38	No Sample Provided
Breath Volume	0.000	03:38	
Air Blank	0.000	03:38	
Air Blank	0.000	03:40	
Subject Sample	# 2	03:41	Subject Test Refused
Breath Volume	0.000	03:41	
Air Blank	0.000	03:41	
Control Test	0.080	03:41	
Air Blank	0.000	03:42	
Diagnostics Check	OK	03:42	

Data Download Date/Time: 09/21/2017 20:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRINSKO,THOMAS,K
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: GUYTON,JOHN
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 08/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:04

SUBJECT'S LAST NAME: STORY

DATE/TIME OF TEST: 09/04/2017 00:32:06

Diagnostics Check	OK	00:24	
Air Blank	0.000	00:25	
Control Test	0.079	00:25	
Air Blank	0.000	00:26	
Subject Sample	# 1 0.246	00:26	
Breath Volume	1.160	00:26	
Air Blank	0.000	00:27	
Air Blank	0.000	00:29	
Subject Sample	# 2 0.229	00:30	
Breath Volume	1.410	00:30	
Air Blank	0.000	00:30	
Control Test	0.079	00:31	
Air Blank		00:31	RFI Detect
Air Blank	0.000	00:32	

Data Download Date/Time: 09/21/2017 20:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEMEN,AARON,M
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: STEMEN,AARON
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 08/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:16

SUBJECT'S LAST NAME: BAHORSKI

DATE/TIME OF TEST: 09/07/2017 01:47:37

Diagnostics Check	OK	01:39	
Air Blank	0.000	01:39	
Control Test	0.079	01:39	
Air Blank	0.000	01:40	
Subject Sample	# 1	01:43	No Sample Provided
Breath Volume	0.000	01:43	
Air Blank	0.000	01:43	
Air Blank	0.000	01:45	
Subject Sample	# 2	01:46	Subject Test Refused
Breath Volume	0.000	01:46	
Air Blank	0.000	01:46	
Control Test	0.079	01:47	
Air Blank	0.000	01:47	
Diagnostics Check	OK	01:47	

Data Download Date/Time: 09/21/2017 20:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VALDES,LYNNE,S
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: ALLEN,LINDSAY
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 08/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: LEE

DATE/TIME OF TEST: 09/13/2017 00:06:58

Diagnostics Check		OK	00:00
Air Blank		0.000	00:00
Control Test		0.078	00:01
Air Blank		0.000	00:01
Subject Sample	# 1	0.209	00:02
Breath Volume		1.644	00:02
Air Blank		0.000	00:02
Air Blank		0.000	00:04
Subject Sample	# 2	0.204	00:05
Breath Volume		2.113	00:05
Air Blank		0.000	00:06
Control Test		0.078	00:06
Air Blank		0.000	00:06
Diagnostics Check		OK	00:06

Data Download Date/Time: 09/21/2017 20:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEPHENS, TRAVIS, L
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: OSBORNE, OWEN
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 08/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 14:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:25

SUBJECT'S LAST NAME: MCGRATH

DATE/TIME OF TEST: 09/15/2017 15:47:10

Diagnostics Check	OK	15:45	
Air Blank	0.000	15:46	
Control Test	0.077	15:46	Interferent Detect
Air Blank	0.000	15:47	

Data Download Date/Time: 09/21/2017 20:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEPHENS, TRAVIS, L
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: OSBORNE, OWEN
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 08/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 14:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:25

SUBJECT'S LAST NAME: MCGRATH

DATE/TIME OF TEST: 09/15/2017 15:57:15

Diagnostics Check	OK	15:50
Air Blank	0.000	15:51
Control Test	0.078	15:51
Air Blank	0.000	15:52
Subject Sample # 1	0.180	15:52
Breath Volume	2.308	15:52
Air Blank	0.000	15:53
Air Blank	0.000	15:55
Subject Sample # 2	0.184	15:55
Breath Volume	1.953	15:55
Air Blank	0.000	15:56
Control Test	0.078	15:56
Air Blank	0.000	15:57
Diagnostics Check	OK	15:57

Data Download Date/Time: 09/21/2017 20:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOPROWSKI,JOHN,J
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: KOPROWSKI,JOHN
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 08/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 17:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:05

SUBJECT'S LAST NAME: WOODBURY

DATE/TIME OF TEST: 09/16/2017 18:38:34

Diagnostics Check	OK	18:27	
Air Blank	0.000	18:27	
Control Test	0.078	18:27	
Air Blank	0.000	18:28	
Subject Sample	# 1	18:31	No Sample Provided
Breath Volume	0.000	18:31	
Air Blank	0.000	18:32	
Air Blank	0.000	18:34	
Subject Sample	# 2	18:37	No Sample Provided
Breath Volume	0.000	18:37	
Air Blank	0.000	18:37	
Control Test	0.079	18:37	
Air Blank	0.000	18:38	
Diagnostics Check	OK	18:38	

Data Download Date/Time: 09/21/2017 20:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEMEN,AARON,M
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: STEMEN,AARON
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 08/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 19:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:24

SUBJECT'S LAST NAME: MORGAN

DATE/TIME OF TEST: 09/16/2017 19:52:49

Diagnostics Check	OK	19:44	
Air Blank	0.000	19:44	
Control Test	0.079	19:45	
Air Blank	0.000	19:45	
Subject Sample	# 1	19:48	No Sample Provided
Breath Volume	0.000	19:48	
Air Blank	0.000	19:49	
Air Blank	0.000	19:51	
Subject Sample	# 2	19:51	Subject Test Refused
Breath Volume	0.000	19:51	
Air Blank	0.000	19:51	
Control Test	0.079	19:52	
Air Blank	0.000	19:52	
Diagnostics Check	OK	19:52	

Data Download Date/Time: 09/21/2017 20:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEMEN,AARON,M
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: STEMEN,AARON
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 08/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:48

SUBJECT'S LAST NAME: REDWINE

DATE/TIME OF TEST: 09/17/2017 04:18:16

Diagnostics Check	OK	04:09	
Air Blank	0.000	04:09	
Control Test	0.079	04:10	
Air Blank	0.000	04:10	
Subject Sample	# 1	04:13	No Sample Provided
Breath Volume	0.000	04:13	
Air Blank	0.000	04:14	
Air Blank	0.000	04:16	
Subject Sample	# 2	04:16	Subject Test Refused
Breath Volume	0.000	04:16	
Air Blank	0.000	04:17	
Control Test	0.079	04:17	
Air Blank	0.000	04:18	
Diagnostics Check	OK	04:18	

Data Download Date/Time: 09/21/2017 20:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOSTICK,JESSE,W
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: BOSTICK,JESSE
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 08/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:12

SUBJECT'S LAST NAME: LOWER

DATE/TIME OF TEST: 09/17/2017 05:40:18

Diagnostics Check	OK	05:33
Air Blank	0.000	05:33
Control Test	0.079	05:34
Air Blank	0.000	05:34
Subject Sample # 1	0.000	05:35
Breath Volume	1.816	05:35
Air Blank	0.000	05:36
Air Blank	0.000	05:38
Subject Sample # 2	0.000	05:38
Breath Volume	1.855	05:38
Air Blank	0.000	05:39
Control Test	0.078	05:39
Air Blank	0.000	05:40
Diagnostics Check	OK	05:40

Data Download Date/Time: 09/21/2017 20:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SOWARDS,JACOB,T
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: GUYTON,JOHN
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 08/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: LEE

DATE/TIME OF TEST: 09/20/2017 01:11:31

Diagnostics Check		OK	01:05
Air Blank		0.000	01:05
Control Test		0.079	01:06
Air Blank		0.000	01:06
Subject Sample	# 1	0.246	01:07
Breath Volume		2.652	01:07
Air Blank		0.000	01:07
Air Blank		0.000	01:09
Subject Sample	# 2	0.262	01:09
Breath Volume		2.277	01:09
Air Blank		0.000	01:10
Control Test		0.078	01:10
Air Blank		0.000	01:11
Diagnostics Check		OK	01:11

Data Download Date/Time: 09/21/2017 20:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEMEN,AARON,M
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: STEMEN,AARON
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 09/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:41

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 09/23/2017 02:10:24

Diagnostics Check	OK	02:04
Air Blank	0.000	02:04
Control Test	0.079	02:05
Air Blank	0.000	02:05
Subject Sample # 1	0.211	02:06
Breath Volume	1.453	02:06
Air Blank	0.000	02:06
Air Blank	0.000	02:08
Subject Sample # 2	0.206	02:08
Breath Volume	1.132	02:08
Air Blank	0.000	02:09
Control Test	0.078	02:09
Air Blank	0.000	02:10
Diagnostics Check	OK	02:10

Data Download Date/Time: 10/13/2017 21:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITE,ROBERT,P
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: WHITE,ROBERT
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 09/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: RIDDELL

DATE/TIME OF TEST: 09/23/2017 02:45:18

Diagnostics Check	OK	02:42	
Air Blank	0.000	02:42	
Control Test	0.079	02:42	
Air Blank	0.000	02:43	
Subject Sample	# 1	02:43	Subject Test Refused
Breath Volume	0.000	02:43	
Air Blank	0.000	02:44	
Control Test	0.079	02:44	
Air Blank	0.000	02:45	
Diagnostics Check	OK	02:45	

Data Download Date/Time: 10/13/2017 21:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEMEN,AARON,M
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: STEMEN,AARON
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 09/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:41

SUBJECT'S LAST NAME: BOLAR

DATE/TIME OF TEST: 09/24/2017 00:16:29

Diagnostics Check	OK	00:05	
Air Blank	0.000	00:05	
Control Test	0.079	00:05	
Air Blank	0.000	00:06	
Subject Sample	# 1	00:09	No Sample Provided
Breath Volume	0.000	00:09	
Air Blank	0.000	00:09	
Air Blank	0.000	00:11	
Subject Sample	# 2	00:15	No Sample Provided
Breath Volume	0.000	00:15	
Air Blank	0.000	00:15	
Control Test	0.080	00:15	
Air Blank	0.000	00:16	
Diagnostics Check	OK	00:16	

Data Download Date/Time: 10/13/2017 21:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOSTICK,JESSE,W
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: QUINN,MATTHEW
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 09/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:49

SUBJECT'S LAST NAME: QUINN

DATE/TIME OF TEST: 09/24/2017 04:15:47

Diagnostics Check	OK	04:09
Air Blank	0.000	04:10
Control Test	0.079	04:10
Air Blank	0.000	04:11
Subject Sample # 1	0.170	04:11
Breath Volume	2.449	04:11
Air Blank	0.000	04:12
Air Blank	0.000	04:13
Subject Sample # 2	0.170	04:14
Breath Volume	1.695	04:14
Air Blank	0.000	04:14
Control Test	0.078	04:15
Air Blank	0.000	04:15
Diagnostics Check	OK	04:15

Data Download Date/Time: 10/13/2017 21:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SOWARDS,JACOB,T
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: GUYTON,JOHN
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 09/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:22

SUBJECT'S LAST NAME: MCNABB

DATE/TIME OF TEST: 09/28/2017 00:56:17

Diagnostics Check		OK	00:42	
Air Blank		0.000	00:42	
Control Test		0.078	00:42	
Air Blank		0.000	00:43	
Subject Sample	# 1	0.092	00:46	Volume Not Met
Breath Volume		0.535	00:46	
Air Blank		0.000	00:47	
Air Blank		0.000	00:49	
Subject Sample	# 2	0.113	00:50	
Breath Volume		1.554	00:50	
Air Blank		0.000	00:51	
Air Blank		0.000	00:53	
Subject Sample	# 3	0.111	00:54	
Breath Volume		1.285	00:54	
Air Blank		0.000	00:55	
Control Test		0.077	00:55	
Air Blank		0.000	00:56	
Diagnostics Check		OK	00:56	

Data Download Date/Time: 10/13/2017 21:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SOWARDS,JACOB,T
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: BOSTICK,ANTHONY
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 09/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:42

SUBJECT'S LAST NAME: MANGIONE

DATE/TIME OF TEST: 09/28/2017 02:05:41

Diagnostics Check	OK	02:02	
Air Blank	0.000	02:03	
Control Test	0.078	02:03	
Air Blank	0.000	02:04	
Subject Sample	# 1	02:04	Subject Test Refused
Breath Volume	0.000	02:04	
Air Blank	0.000	02:04	
Control Test	0.079	02:05	
Air Blank	0.000	02:05	
Diagnostics Check	OK	02:05	

Data Download Date/Time: 10/13/2017 21:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEMEN,AARON,M
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: STEMEN,AARON
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 09/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:47

SUBJECT'S LAST NAME: DIEGIDIO

DATE/TIME OF TEST: 09/29/2017 04:17:26

Diagnostics Check	OK	04:10
Air Blank	0.000	04:11
Control Test	0.079	04:11
Air Blank	0.000	04:11
Subject Sample # 1	0.191	04:12
Breath Volume	2.171	04:12
Air Blank	0.000	04:13
Air Blank	0.000	04:15
Subject Sample # 2	0.182	04:15
Breath Volume	1.316	04:15
Air Blank	0.000	04:16
Control Test	0.079	04:16
Air Blank	0.000	04:17
Diagnostics Check	OK	04:17

Data Download Date/Time: 10/13/2017 21:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,CALEB,N
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ELLER,ASHLEY
ARRESTING OFFICER AGENCY: Apopka PD
LAST AGENCY INSPECTION: 08/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: TURSINI

DATE/TIME OF TEST: 09/01/2017 03:15:56

Diagnostics Check		OK	03:00	
Air Blank		0.000	03:00	
Control Test		0.081	03:00	
Air Blank		0.000	03:01	
Subject Sample	# 1	0.212	03:03	
Breath Volume		1.273	03:03	
Air Blank		0.000	03:03	
Air Blank		0.000	03:05	
Subject Sample	# 2	0.193	03:08	Volume Not Met
Breath Volume		0.156	03:08	
Air Blank		0.000	03:09	
Air Blank		0.000	03:11	
Subject Sample	# 3	0.153	03:14	Volume Not Met
Breath Volume		0.351	03:14	
Air Blank		0.000	03:15	
Control Test		0.079	03:15	
Air Blank		0.000	03:15	
Diagnostics Check		OK	03:15	

Data Download Date/Time: 09/07/2017 16:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,COURTNEY,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MASSICOTTE,MICHAEL
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 08/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:24

SUBJECT'S LAST NAME: LEWIS

DATE/TIME OF TEST: 09/01/2017 03:55:47

Diagnostics Check	OK	03:48
Air Blank	0.000	03:49
Control Test	0.080	03:49
Air Blank	0.000	03:50
Subject Sample # 1	0.160	03:50
Breath Volume	1.828	03:50
Air Blank	0.000	03:51
Air Blank	0.000	03:53
Subject Sample # 2	0.157	03:54
Breath Volume	1.886	03:54
Air Blank	0.000	03:54
Control Test	0.081	03:55
Air Blank	0.000	03:55
Diagnostics Check	OK	03:55

Data Download Date/Time: 09/07/2017 16:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELBON,NICHOLAS,C
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HARRIS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 08/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:31

SUBJECT'S LAST NAME: DICKSON

DATE/TIME OF TEST: 09/01/2017 05:03:18

Diagnostics Check	OK	04:57
Air Blank	0.000	04:57
Control Test	0.080	04:57
Air Blank	0.000	04:58
Subject Sample # 1	0.159	04:58
Breath Volume	3.109	04:58
Air Blank	0.000	04:59
Air Blank	0.000	05:01
Subject Sample # 2	0.162	05:01
Breath Volume	3.175	05:01
Air Blank	0.000	05:02
Control Test	0.080	05:02
Air Blank	0.000	05:03
Diagnostics Check	OK	05:03

Data Download Date/Time: 09/07/2017 16:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DRACH,STEFAN
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 08/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:28

SUBJECT'S LAST NAME: KURASTOLER

DATE/TIME OF TEST: 09/01/2017 23:59:39

Diagnostics Check	OK	23:52
Air Blank	0.000	23:53
Control Test	0.080	23:53
Air Blank	0.000	23:53
Subject Sample # 1	0.093	23:54
Breath Volume	1.562	23:54
Air Blank	0.000	23:55
Air Blank	0.000	23:57
Subject Sample # 2	0.088	23:58
Breath Volume	2.335	23:58
Air Blank	0.000	23:58
Control Test	0.081	23:59
Air Blank	0.000	23:59
Diagnostics Check	OK	23:59

Data Download Date/Time: 09/07/2017 16:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,CALEB,N
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HAWKINS,DEBORAH
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 08/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:17

SUBJECT'S LAST NAME: IRWIN

DATE/TIME OF TEST: 09/02/2017 01:50:55

Diagnostics Check	OK	01:44
Air Blank	0.000	01:44
Control Test	0.080	01:44
Air Blank	0.000	01:45
Subject Sample # 1	0.268	01:46
Breath Volume	3.039	01:46
Air Blank	0.000	01:46
Air Blank	0.000	01:48
Subject Sample # 2	0.282	01:49
Breath Volume	3.332	01:49
Air Blank	0.000	01:50
Control Test	0.079	01:50
Air Blank	0.000	01:50
Diagnostics Check	OK	01:50

Data Download Date/Time: 09/07/2017 16:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,COURTNEY,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CSIZMADIA,ANDRAS
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 08/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:36

SUBJECT'S LAST NAME: VALLADARES DIAZ

DATE/TIME OF TEST: 09/02/2017 02:18:18

Diagnostics Check		OK	02:04	
Air Blank		0.000	02:05	
Control Test		0.080	02:05	
Air Blank		0.000	02:05	
Subject Sample	# 1	0.000	02:09	Volume Not Met
Breath Volume		0.351	02:09	
Air Blank		0.000	02:09	
Air Blank		0.000	02:11	
Subject Sample	# 2	0.000	02:12	
Breath Volume		1.664	02:12	
Air Blank		0.000	02:12	
Air Blank		0.000	02:14	
Subject Sample	# 3	0.000	02:16	
Breath Volume		1.453	02:16	
Air Blank		0.000	02:17	
Control Test		0.080	02:17	
Air Blank		0.000	02:18	
Diagnostics Check		OK	02:18	

Data Download Date/Time: 09/07/2017 16:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,CALEB,N
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: LUGO,OSCAR
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 08/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: HUTCHISON

DATE/TIME OF TEST: 09/02/2017 04:05:15

Diagnostics Check	OK	03:54
Air Blank	0.000	03:55
Control Test	0.082	03:55
Air Blank	0.000	03:55
Subject Sample # 1	0.202	03:58
Breath Volume	1.527	03:58
Air Blank	0.000	03:58
Air Blank	0.000	04:00
Subject Sample # 2	0.192	04:03
Breath Volume	1.359	04:03
Air Blank	0.000	04:04
Control Test	0.080	04:04
Air Blank	0.000	04:05
Diagnostics Check	OK	04:05

Data Download Date/Time: 09/07/2017 16:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELBON,NICHOLAS,C
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: THOMPSON,JOSEPH
ARRESTING OFFICER AGENCY: Winter Garden PD
LAST AGENCY INSPECTION: 08/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:02

SUBJECT'S LAST NAME: HINES

DATE/TIME OF TEST: 09/03/2017 00:33:01

Diagnostics Check		OK	00:26
Air Blank		0.000	00:27
Control Test		0.081	00:27
Air Blank		0.000	00:27
Subject Sample	# 1	0.046	00:28
Breath Volume		1.882	00:28
Air Blank		0.000	00:28
Air Blank		0.000	00:30
Subject Sample	# 2	0.041	00:31
Breath Volume		1.687	00:31
Air Blank		0.000	00:32
Control Test		0.081	00:32
Air Blank		0.000	00:32
Diagnostics Check		OK	00:33

Data Download Date/Time: 09/07/2017 16:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HAWKINS,DEB
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 08/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:53

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 09/03/2017 01:23:10

Diagnostics Check	OK	01:16
Air Blank	0.000	01:17
Control Test	0.081	01:17
Air Blank	0.000	01:18
Subject Sample # 1	0.191	01:18
Breath Volume	1.375	01:18
Air Blank	0.000	01:19
Air Blank	0.000	01:21
Subject Sample # 2	0.187	01:21
Breath Volume	1.175	01:21
Air Blank	0.000	01:22
Control Test	0.080	01:22
Air Blank	0.000	01:23
Diagnostics Check	OK	01:23

Data Download Date/Time: 09/07/2017 16:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ORTIZ, ROBERTO
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 08/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:27

SUBJECT'S LAST NAME: CHAVEZ PEREZ

DATE/TIME OF TEST: 09/03/2017 02:58:47

Diagnostics Check	OK	02:52
Air Blank	0.000	02:53
Control Test	0.080	02:53
Air Blank	0.000	02:54
Subject Sample # 1	0.099	02:54
Breath Volume	2.394	02:54
Air Blank	0.000	02:55
Air Blank	0.000	02:57
Subject Sample # 2	0.092	02:57
Breath Volume	1.890	02:57
Air Blank	0.000	02:57
Control Test	0.080	02:58
Air Blank	0.000	02:58
Diagnostics Check	OK	02:58

Data Download Date/Time: 09/07/2017 16:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HAWKINS,DEB
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 08/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:31

SUBJECT'S LAST NAME: LEE

DATE/TIME OF TEST: 09/03/2017 05:04:20

Diagnostics Check	OK	04:55
Air Blank	0.000	04:55
Control Test	0.081	04:55
Air Blank	0.000	04:56
Subject Sample # 1	0.195	04:58
Breath Volume	1.363	04:58
Air Blank	0.000	04:59
Air Blank	0.000	05:01
Subject Sample # 2	0.184	05:02
Breath Volume	2.027	05:02
Air Blank	0.000	05:03
Control Test	0.080	05:03
Air Blank	0.000	05:04
Diagnostics Check	OK	05:04

Data Download Date/Time: 09/07/2017 16:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: RUSH,BRANDON
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 08/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 11:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:45

SUBJECT'S LAST NAME: ROCHA

DATE/TIME OF TEST: 09/04/2017 12:18:49

Diagnostics Check	OK	12:12
Air Blank	0.000	12:12
Control Test	0.081	12:13
Air Blank	0.000	12:13
Subject Sample # 1	0.166	12:14
Breath Volume	2.992	12:14
Air Blank	0.000	12:14
Air Blank	0.000	12:16
Subject Sample # 2	0.159	12:17
Breath Volume	1.890	12:17
Air Blank	0.000	12:17
Control Test	0.080	12:18
Air Blank	0.000	12:18
Diagnostics Check	OK	12:18

Data Download Date/Time: 09/07/2017 16:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OKEEFE,ALEXIS,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HAWKINS,D
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 08/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: MARDONES FIGUEROA

DATE/TIME OF TEST: 09/07/2017 01:46:44

Diagnostics Check		OK	01:33	
Air Blank		0.000	01:33	
Control Test		0.081	01:34	
Air Blank		0.000	01:34	
Subject Sample	# 1	0.126	01:37	Volume Not Met
Breath Volume		0.609	01:37	
Air Blank		0.000	01:38	
Air Blank		0.000	01:40	
Subject Sample	# 2	0.139	01:42	
Breath Volume		1.722	01:42	
Air Blank		0.000	01:42	
Air Blank		0.000	01:44	
Subject Sample	# 3	0.149	01:45	
Breath Volume		1.851	01:45	
Air Blank		0.000	01:45	
Control Test		0.080	01:46	
Air Blank		0.000	01:46	
Diagnostics Check		OK	01:46	

Data Download Date/Time: 09/07/2017 16:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,CALEB,N
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: LOWE,KEVIN
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 08/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:51

SUBJECT'S LAST NAME: HARRIPAUL

DATE/TIME OF TEST: 09/07/2017 05:28:39

Diagnostics Check		OK	05:14	
Air Blank		0.000	05:15	
Control Test		0.081	05:15	
Air Blank		0.000	05:16	
Subject Sample	# 1	0.156	05:19	Volume Not Met
Breath Volume		0.746	05:19	
Air Blank		0.000	05:19	
Air Blank		0.000	05:21	
Subject Sample	# 2	0.176	05:22	
Breath Volume		1.492	05:22	
Air Blank		0.000	05:22	
Air Blank		0.000	05:24	
Subject Sample	# 3	0.164	05:27	
Breath Volume		1.921	05:27	
Air Blank		0.000	05:27	
Control Test		0.080	05:28	
Air Blank		0.000	05:28	
Diagnostics Check		OK	05:28	

Data Download Date/Time: 09/07/2017 16:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ, MICHAEL, J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CORTES, ADAM
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 08/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 13:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:26

SUBJECT'S LAST NAME: SKIPPINGS

DATE/TIME OF TEST: 09/08/2017 15:06:47

Diagnostics Check	OK	15:00
Air Blank	0.000	15:00
Control Test	0.082	15:00
Air Blank	0.000	15:01
Subject Sample # 1	0.000	15:02
Breath Volume	1.750	15:02
Air Blank	0.000	15:02
Air Blank	0.000	15:05
Subject Sample # 2	0.000	15:05
Breath Volume	1.609	15:05
Air Blank	0.000	15:05
Control Test	0.082	15:06
Air Blank	0.000	15:06
Diagnostics Check	OK	15:06

Data Download Date/Time: 09/14/2017 12:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,CALEB,N
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BRENARD,KYLE
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 08/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 09/09/2017 01:23:23

Diagnostics Check	OK	01:16
Air Blank	0.000	01:17
Control Test	0.081	01:17
Air Blank	0.000	01:18
Subject Sample # 1	0.169	01:18
Breath Volume	1.386	01:18
Air Blank	0.000	01:19
Air Blank	0.000	01:21
Subject Sample # 2	0.168	01:21
Breath Volume	1.859	01:21
Air Blank	0.000	01:22
Control Test	0.081	01:22
Air Blank	0.000	01:23
Diagnostics Check	OK	01:23

Data Download Date/Time: 09/14/2017 12:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELBON,NICHOLAS,C
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DRACH,STEFAN
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 08/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:53

SUBJECT'S LAST NAME: ZOTO

DATE/TIME OF TEST: 09/09/2017 03:25:01

Diagnostics Check	OK	03:18
Air Blank	0.000	03:19
Control Test	0.081	03:19
Air Blank	0.000	03:20
Subject Sample # 1	0.198	03:20
Breath Volume	3.378	03:20
Air Blank	0.000	03:21
Air Blank	0.000	03:23
Subject Sample # 2	0.197	03:23
Breath Volume	3.855	03:23
Air Blank	0.000	03:24
Control Test	0.081	03:24
Air Blank	0.000	03:24
Diagnostics Check	OK	03:25

Data Download Date/Time: 09/14/2017 12:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELBON,NICHOLAS',C
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HALL,CATHERINE
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 08/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: HABERMAN

DATE/TIME OF TEST: 09/09/2017 04:09:42

Diagnostics Check	OK	04:03
Air Blank	0.000	04:04
Control Test	0.081	04:04
Air Blank	0.000	04:04
Subject Sample # 1	0.163	04:05
Breath Volume	1.253	04:05
Air Blank	0.000	04:05
Air Blank	0.000	04:07
Subject Sample # 2	0.171	04:08
Breath Volume	3.539	04:08
Air Blank	0.000	04:08
Control Test	0.081	04:09
Air Blank	0.000	04:09
Diagnostics Check	OK	04:09

Data Download Date/Time: 09/14/2017 12:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ, MICHAEL, J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: AZEVEDO, FABIO
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 08/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 05:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:41

SUBJECT'S LAST NAME: HERNANDEZ GINES

DATE/TIME OF TEST: 09/09/2017 06:11:03

Diagnostics Check		OK	06:04
Air Blank		0.000	06:05
Control Test		0.081	06:05
Air Blank		0.000	06:06
Subject Sample	# 1	0.188	06:06
Breath Volume		2.660	06:06
Air Blank		0.000	06:07
Air Blank		0.000	06:09
Subject Sample	# 2	0.188	06:09
Breath Volume		2.804	06:09
Air Blank		0.000	06:10
Control Test		0.081	06:10
Air Blank		0.000	06:10
Diagnostics Check		OK	06:11

Data Download Date/Time: 09/14/2017 12:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OKEEFE,ALEXIS,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: EDWARDS,JOHN
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 08/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:25

SUBJECT'S LAST NAME: BIGGINS

DATE/TIME OF TEST: 09/11/2017 22:54:53

Diagnostics Check	OK	22:48
Air Blank	0.000	22:48
Control Test	0.081	22:48
Air Blank	0.000	22:49
Subject Sample # 1	0.000	22:50
Breath Volume	1.699	22:50
Air Blank	0.000	22:50
Air Blank	0.000	22:52
Subject Sample # 2	0.000	22:53
Breath Volume	1.960	22:53
Air Blank	0.000	22:53
Control Test	0.081	22:54
Air Blank	0.000	22:54
Diagnostics Check	OK	22:54

Data Download Date/Time: 09/14/2017 12:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OKEEFE,ALEXIS,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: WEAVER,W
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 08/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: HORNEDO SILVA

DATE/TIME OF TEST: 09/12/2017 03:30:50

Diagnostics Check	OK	03:21
Air Blank	0.000	03:22
Control Test	0.081	03:22
Air Blank	0.000	03:22
Subject Sample # 1	0.280	03:23
Breath Volume	2.714	03:23
Air Blank	0.000	03:23
Air Blank	0.000	03:25
Subject Sample # 2	0.255	03:26
Breath Volume	1.582	03:26
Air Blank	0.000	03:26
Air Blank	0.000	03:28
Subject Sample # 3	0.267	03:29
Breath Volume	2.324	03:29
Air Blank	0.000	03:29
Control Test	0.081	03:30
Air Blank	0.000	03:30
Diagnostics Check	OK	03:30

Data Download Date/Time: 09/14/2017 12:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OKEEFE,ALEXIS,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HALL,C
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 08/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:45

SUBJECT'S LAST NAME: KUMAR

DATE/TIME OF TEST: 09/12/2017 05:14:18

Diagnostics Check	OK	05:07
Air Blank	0.000	05:07
Control Test	0.081	05:08
Air Blank	0.000	05:08
Subject Sample # 1	0.093	05:09
Breath Volume	1.667	05:09
Air Blank	0.000	05:10
Air Blank	0.000	05:12
Subject Sample # 2	0.095	05:12
Breath Volume	2.101	05:12
Air Blank	0.000	05:13
Control Test	0.081	05:13
Air Blank	0.000	05:14
Diagnostics Check	OK	05:14

Data Download Date/Time: 09/14/2017 12:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OKEEFE,ALEXIS,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: WATTS,ANTHONY
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 08/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 18:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:00

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 09/12/2017 19:37:17

Diagnostics Check	OK	19:28	
Air Blank	0.000	19:29	
Control Test	0.081	19:29	
Air Blank	0.000	19:30	
Subject Sample	# 1 0.161	19:33	Volume Not Met
Breath Volume	0.699	19:33	
Air Blank	0.000	19:33	
Air Blank	0.000	19:35	
Subject Sample	# 2	19:35	Subject Test Refused
Breath Volume	0.000	19:35	
Air Blank	0.000	19:36	
Control Test	0.078	19:36	
Air Blank	0.000	19:37	
Diagnostics Check	OK	19:37	

Data Download Date/Time: 09/14/2017 12:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,CALEB,N
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HAWKINS,DEBORAH
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 08/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:17

SUBJECT'S LAST NAME: SERRANO

DATE/TIME OF TEST: 09/12/2017 23:47:00

Diagnostics Check	OK	23:40
Air Blank	0.000	23:41
Control Test	0.081	23:41
Air Blank	0.000	23:42
Subject Sample # 1	0.174	23:42
Breath Volume	1.246	23:42
Air Blank	0.000	23:43
Air Blank	0.000	23:45
Subject Sample # 2	0.174	23:45
Breath Volume	2.304	23:45
Air Blank	0.000	23:46
Control Test	0.081	23:46
Air Blank	0.000	23:46
Diagnostics Check	OK	23:47

Data Download Date/Time: 09/14/2017 12:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OKEEFE,ALEXIS,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DRACH,S
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 08/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: HABEGGER

DATE/TIME OF TEST: 09/13/2017 03:38:19

Diagnostics Check	OK	03:32
Air Blank	0.000	03:32
Control Test	0.081	03:33
Air Blank	0.000	03:33
Subject Sample # 1	0.196	03:34
Breath Volume	2.136	03:34
Air Blank	0.000	03:34
Air Blank	0.000	03:36
Subject Sample # 2	0.177	03:36
Breath Volume	1.250	03:36
Air Blank	0.000	03:37
Control Test	0.081	03:37
Air Blank	0.000	03:38
Diagnostics Check	OK	03:38

Data Download Date/Time: 09/14/2017 12:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,COURTNEY,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HAWKINS,DEB
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 08/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:53

SUBJECT'S LAST NAME: MONTES

DATE/TIME OF TEST: 09/13/2017 04:22:07

Diagnostics Check	OK	04:15
Air Blank	0.000	04:16
Control Test	0.080	04:16
Air Blank	0.000	04:17
Subject Sample # 1	0.110	04:17
Breath Volume	1.566	04:17
Air Blank	0.000	04:18
Air Blank	0.000	04:20
Subject Sample # 2	0.109	04:20
Breath Volume	1.808	04:20
Air Blank	0.000	04:21
Control Test	0.081	04:21
Air Blank	0.000	04:22
Diagnostics Check	OK	04:22

Data Download Date/Time: 09/14/2017 12:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OKEEFE,ALEXIS,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ROBBINS,N
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 08/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:25

SUBJECT'S LAST NAME: WHITELEY

DATE/TIME OF TEST: 09/13/2017 21:54:09

Diagnostics Check	OK	21:48
Air Blank	0.000	21:48
Control Test	0.082	21:49
Air Blank	0.000	21:49
Subject Sample # 1	0.000	21:49
Breath Volume	1.636	21:49
Air Blank	0.000	21:50
Air Blank	0.000	21:52
Subject Sample # 2	0.000	21:52
Breath Volume	2.617	21:52
Air Blank	0.000	21:53
Control Test	0.081	21:53
Air Blank	0.000	21:54
Diagnostics Check	OK	21:54

Data Download Date/Time: 09/14/2017 12:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,CALEB,N
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DRACH,STEFAN
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 08/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:47

SUBJECT'S LAST NAME: GIRAUDMONSANTO

DATE/TIME OF TEST: 09/13/2017 23:28:03

Diagnostics Check		OK	23:11	
Air Blank		0.000	23:11	
Control Test		0.081	23:11	
Air Blank		0.000	23:12	
Subject Sample	# 1	0.181	23:15	Volume Not Met
Breath Volume		0.726	23:15	
Air Blank		0.000	23:16	
Air Blank		0.000	23:18	
Subject Sample	# 2	0.182	23:20	
Breath Volume		1.148	23:20	
Air Blank		0.000	23:21	
Air Blank		0.000	23:23	
Subject Sample	# 3	0.172	23:26	Slope Not Level
Breath Volume		1.191	23:26	
Air Blank		0.000	23:27	
Control Test		0.079	23:27	
Air Blank		0.000	23:27	
Diagnostics Check		OK	23:28	

Data Download Date/Time: 09/14/2017 12:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON, CALEB, N
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DRACH, STEFAN
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 08/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:47

SUBJECT'S LAST NAME: GIRAUD MONSANTO

DATE/TIME OF TEST: 09/13/2017 23:38:14

Diagnostics Check	OK	23:29	
Air Blank	0.000	23:30	
Control Test	0.080	23:30	
Air Blank	0.000	23:31	
Subject Sample	# 1 0.169	23:34	Volume Not Met
Breath Volume	0.871	23:34	
Air Blank	0.000	23:34	
Air Blank	0.000	23:36	
Subject Sample	# 2	23:36	Subject Test Refused
Breath Volume	0.000	23:36	
Air Blank	0.000	23:37	
Control Test	0.078	23:37	
Air Blank	0.000	23:38	
Diagnostics Check	OK	23:38	

Data Download Date/Time: 09/14/2017 12:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,CALEB,N
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MCWHERTER,PETER
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 08/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:04

SUBJECT'S LAST NAME: MALLICK

DATE/TIME OF TEST: 09/14/2017 01:33:23

Diagnostics Check	OK	01:26
Air Blank	0.000	01:26
Control Test	0.080	01:27
Air Blank	0.000	01:27
Subject Sample # 1	0.145	01:28
Breath Volume	1.769	01:28
Air Blank	0.000	01:29
Air Blank	0.000	01:31
Subject Sample # 2	0.147	01:31
Breath Volume	1.410	01:31
Air Blank	0.000	01:32
Control Test	0.080	01:32
Air Blank	0.000	01:33
Diagnostics Check	OK	01:33

Data Download Date/Time: 09/14/2017 12:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OKEEFE,ALEXIS,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HAWKINS,D
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 08/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: RUIZ

DATE/TIME OF TEST: 09/14/2017 03:34:37

Diagnostics Check	OK	03:26
Air Blank	0.000	03:27
Control Test	0.081	03:27
Air Blank	0.000	03:27
Subject Sample # 1	0.195	03:29
Breath Volume	1.437	03:29
Air Blank	0.000	03:29
Air Blank	0.000	03:31
Subject Sample # 2	0.195	03:33
Breath Volume	1.937	03:33
Air Blank	0.000	03:33
Control Test	0.080	03:34
Air Blank	0.000	03:34
Diagnostics Check	OK	03:34

Data Download Date/Time: 09/14/2017 12:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: VAUGHN,JACOB
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 08/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:07

SUBJECT'S LAST NAME: PINCHINAT

DATE/TIME OF TEST: 09/14/2017 04:42:16

Diagnostics Check		OK	04:30	
Air Blank		0.000	04:31	
Control Test		0.081	04:31	
Air Blank		0.000	04:31	
Subject Sample	# 1	0.151	04:35	Volume Not Met
Breath Volume		0.347	04:35	
Air Blank		0.000	04:35	
Air Blank		0.000	04:37	
Subject Sample	# 2	0.135	04:40	Volume Not Met
Breath Volume		0.347	04:40	
Air Blank		0.000	04:41	
Control Test		0.080	04:41	
Air Blank		0.000	04:42	
Diagnostics Check		OK	04:42	

Data Download Date/Time: 09/14/2017 12:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: VAUGHN,JACOB
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 08/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:07

SUBJECT'S LAST NAME: PINCHINAT

DATE/TIME OF TEST: 09/14/2017 04:58:26

Diagnostics Check		OK	04:43	
Air Blank		0.000	04:44	
Control Test		0.080	04:44	
Air Blank		0.000	04:44	
Subject Sample	# 1	0.156	04:48	Volume Not Met
Breath Volume		0.812	04:48	
Air Blank		0.000	04:48	
Air Blank		0.000	04:50	
Subject Sample	# 2	0.157	04:51	
Breath Volume		1.277	04:51	
Air Blank		0.000	04:51	
Air Blank		0.000	04:53	
Subject Sample	# 3	0.157	04:56	Volume Not Met
Breath Volume		0.191	04:56	
Air Blank		0.000	04:57	
Control Test		0.080	04:57	
Air Blank		0.000	04:58	
Diagnostics Check		OK	04:58	

Data Download Date/Time: 09/14/2017 12:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OKEEFE,ALEXIS,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: PARKHURST,TIM
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 08/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 17:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:00

SUBJECT'S LAST NAME: PEREA

DATE/TIME OF TEST: 09/14/2017 18:37:40

Diagnostics Check	OK	18:31
Air Blank	0.000	18:32
Control Test	0.082	18:32
Air Blank	0.000	18:32
Subject Sample # 1	0.176	18:33
Breath Volume	1.652	18:33
Air Blank	0.000	18:33
Air Blank	0.000	18:35
Subject Sample # 2	0.183	18:36
Breath Volume	1.441	18:36
Air Blank	0.000	18:36
Control Test	0.080	18:37
Air Blank	0.000	18:37
Diagnostics Check	OK	18:37

Data Download Date/Time: 09/21/2017 10:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HAWKINS,DEB
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 08/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: FAJARDO

DATE/TIME OF TEST: 09/15/2017 02:34:48

Diagnostics Check	OK	02:27
Air Blank	0.000	02:27
Control Test	0.081	02:27
Air Blank	0.000	02:28
Subject Sample # 1	0.070	02:29
Breath Volume	1.523	02:29
Air Blank	0.000	02:30
Air Blank	0.000	02:31
Subject Sample # 2	0.062	02:33
Breath Volume	1.507	02:33
Air Blank	0.000	02:33
Control Test	0.081	02:34
Air Blank	0.000	02:34
Diagnostics Check	OK	02:34

Data Download Date/Time: 09/21/2017 10:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SIMPSON,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 08/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:41

SUBJECT'S LAST NAME: KLEEMAN

DATE/TIME OF TEST: 09/16/2017 00:20:37

Diagnostics Check		OK	00:07	
Air Blank		0.000	00:08	
Control Test		0.081	00:08	
Air Blank		0.000	00:08	
Subject Sample	# 1	0.000	00:12	Volume Not Met
Breath Volume		0.164	00:12	
Air Blank		0.000	00:12	
Air Blank		0.000	00:14	
Subject Sample	# 2	0.000	00:15	
Breath Volume		1.550	00:15	
Air Blank		0.000	00:15	
Air Blank		0.000	00:17	
Subject Sample	# 3	0.000	00:19	
Breath Volume		1.523	00:19	
Air Blank		0.000	00:19	
Control Test		0.081	00:20	
Air Blank		0.000	00:20	
Diagnostics Check		OK	00:20	

Data Download Date/Time: 09/21/2017 10:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,COURTNEY,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MORRIS,STEVE
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 08/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:06

SUBJECT'S LAST NAME: MORGAN

DATE/TIME OF TEST: 09/16/2017 01:36:56

Diagnostics Check		OK	01:30
Air Blank		0.000	01:31
Control Test		0.082	01:31
Air Blank		0.000	01:32
Subject Sample	# 1	0.166	01:32
Breath Volume		2.273	01:32
Air Blank		0.000	01:33
Air Blank		0.000	01:34
Subject Sample	# 2	0.162	01:35
Breath Volume		2.179	01:35
Air Blank		0.000	01:36
Control Test		0.080	01:36
Air Blank		0.000	01:36
Diagnostics Check		OK	01:36

Data Download Date/Time: 09/21/2017 10:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: RUSSAK,KEVIN
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 08/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:43

SUBJECT'S LAST NAME: FLORES

DATE/TIME OF TEST: 09/16/2017 04:15:02

Diagnostics Check	OK	04:09
Air Blank	0.000	04:09
Control Test	0.082	04:09
Air Blank	0.000	04:10
Subject Sample # 1	0.180	04:10
Breath Volume	2.449	04:10
Air Blank	0.000	04:11
Air Blank	0.000	04:13
Subject Sample # 2	0.183	04:13
Breath Volume	3.003	04:13
Air Blank	0.000	04:14
Control Test	0.081	04:14
Air Blank	0.000	04:14
Diagnostics Check	OK	04:15

Data Download Date/Time: 09/21/2017 10:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: KERN,ETHAN
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 08/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:53

SUBJECT'S LAST NAME: OKPUKPAN

DATE/TIME OF TEST: 09/16/2017 05:23:37

Diagnostics Check	OK	05:17
Air Blank	0.000	05:17
Control Test	0.082	05:17
Air Blank	0.000	05:18
Subject Sample # 1	0.186	05:19
Breath Volume	2.003	05:19
Air Blank	0.000	05:19
Air Blank	0.000	05:21
Subject Sample # 2	0.181	05:22
Breath Volume	1.867	05:22
Air Blank	0.000	05:22
Control Test	0.081	05:23
Air Blank	0.000	05:23
Diagnostics Check	OK	05:23

Data Download Date/Time: 09/21/2017 10:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: GAUTAM,ROBERT
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 08/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:46

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 09/17/2017 03:17:25

Diagnostics Check	OK	03:09
Air Blank	0.000	03:10
Control Test	0.081	03:10
Air Blank	0.000	03:11
Subject Sample # 1	0.159	03:12
Breath Volume	1.179	03:12
Air Blank	0.000	03:13
Air Blank	0.000	03:15
Subject Sample # 2	0.170	03:15
Breath Volume	1.203	03:15
Air Blank	0.000	03:16
Control Test	0.080	03:16
Air Blank	0.000	03:17
Diagnostics Check	OK	03:17

Data Download Date/Time: 09/21/2017 10:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CANER,OSVALDO,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: GAUTAM,ROBERT
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 08/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:26

SUBJECT'S LAST NAME: HARDY

DATE/TIME OF TEST: 09/18/2017 04:02:02

Diagnostics Check	OK	03:54
Air Blank	0.000	03:54
Control Test	0.082	03:54
Air Blank	0.000	03:55
Subject Sample # 1	0.056	03:57
Breath Volume	2.109	03:57
Air Blank	0.000	03:58
Air Blank	0.000	04:00
Subject Sample # 2	0.050	04:00
Breath Volume	1.546	04:00
Air Blank	0.000	04:01
Control Test	0.081	04:01
Air Blank	0.000	04:01
Diagnostics Check	OK	04:02

Data Download Date/Time: 09/21/2017 10:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OKEEFE,ALEXIS,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: AGOSTO,A
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 08/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:55

SUBJECT'S LAST NAME: CORREIA

DATE/TIME OF TEST: 09/19/2017 23:28:58

Diagnostics Check	OK	23:20
Air Blank	0.000	23:20
Control Test	0.081	23:20
Air Blank	0.000	23:21
Subject Sample # 1	0.144	23:24
Breath Volume	1.734	23:24
Air Blank	0.000	23:24
Air Blank	0.000	23:26
Subject Sample # 2	0.137	23:27
Breath Volume	1.632	23:27
Air Blank	0.000	23:28
Control Test	0.080	23:28
Air Blank	0.000	23:28
Diagnostics Check	OK	23:28

Data Download Date/Time: 09/21/2017 10:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,CALEB,N
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DONAKOWSKI,JUSTIN
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 08/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:23

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 09/20/2017 01:52:30

Diagnostics Check	OK	01:46
Air Blank	0.000	01:47
Control Test	0.082	01:47
Air Blank	0.000	01:47
Subject Sample # 1	0.000	01:48
Breath Volume	1.343	01:48
Air Blank	0.000	01:48
Air Blank	0.000	01:50
Subject Sample # 2	0.000	01:51
Breath Volume	1.542	01:51
Air Blank	0.000	01:51
Control Test	0.082	01:51
Air Blank	0.000	01:52
Diagnostics Check	OK	01:52

Data Download Date/Time: 09/21/2017 10:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,CALEB,N
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HATCH,WILLIAM
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 08/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:12

SUBJECT'S LAST NAME: PADILLA

DATE/TIME OF TEST: 09/20/2017 05:47:43

Diagnostics Check		OK	05:36	
Air Blank		0.000	05:36	
Control Test		0.082	05:36	
Air Blank		0.000	05:37	
Subject Sample	# 1	0.108	05:40	Volume Not Met
Breath Volume		0.308	05:40	
Air Blank		0.000	05:41	
Air Blank		0.000	05:43	
Subject Sample	# 2	0.109	05:46	Volume Not Met
Breath Volume		0.097	05:46	
Air Blank		0.000	05:46	
Control Test		0.082	05:47	
Air Blank		0.000	05:47	
Diagnostics Check		OK	05:47	

Data Download Date/Time: 09/21/2017 10:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DRACH,STEFAN
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 08/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:31

SUBJECT'S LAST NAME: YOUNGS

DATE/TIME OF TEST: 09/21/2017 00:02:09

Diagnostics Check	OK	23:56
Air Blank	0.000	23:56
Control Test	0.082	23:56
Air Blank	0.000	23:57
Subject Sample # 2	0.000	23:57
Breath Volume	3.804	23:57
Air Blank	0.000	23:58
Air Blank	0.000	00:00
Subject Sample # 1	0.000	00:00
Breath Volume	3.996	00:00
Air Blank	0.000	00:01
Control Test	0.082	00:01
Air Blank	0.000	00:02
Diagnostics Check	OK	00:02

Data Download Date/Time: 09/21/2017 10:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDEN,WILLIAM,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: VINER,DENNIS
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 09/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:23

SUBJECT'S LAST NAME: MATOS-HIDALGO

DATE/TIME OF TEST: 09/23/2017 03:52:00

Diagnostics Check	OK	03:45
Air Blank	0.000	03:45
Control Test	0.081	03:46
Air Blank	0.000	03:46
Subject Sample # 1	0.202	03:47
Breath Volume	3.355	03:47
Air Blank	0.000	03:47
Air Blank	0.000	03:49
Subject Sample # 2	0.204	03:50
Breath Volume	3.281	03:50
Air Blank	0.000	03:51
Control Test	0.081	03:51
Air Blank	0.000	03:51
Diagnostics Check	OK	03:52

Data Download Date/Time: 09/28/2017 13:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: AZEVEDO,FABIO
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 09/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:53

SUBJECT'S LAST NAME: SPINOGATTI

DATE/TIME OF TEST: 09/24/2017 01:22:25

Diagnostics Check		OK	01:15
Air Blank		0.000	01:16
Control Test		0.082	01:16
Air Blank		0.000	01:16
Subject Sample	# 1	0.153	01:17
Breath Volume		1.953	01:17
Air Blank		0.000	01:18
Air Blank		0.000	01:20
Subject Sample	# 2	0.157	01:20
Breath Volume		2.273	01:20
Air Blank		0.000	01:21
Control Test		0.081	01:21
Air Blank		0.000	01:22
Diagnostics Check		OK	01:22

Data Download Date/Time: 09/28/2017 13:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDEN,WILLIAM,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: RIVERA,LIBORIO
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 09/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: LOPEZ XOLO

DATE/TIME OF TEST: 09/25/2017 00:15:06

Diagnostics Check		OK	00:06	
Air Blank		0.000	00:06	
Control Test		0.081	00:06	
Air Blank		0.000	00:07	
Subject Sample	# 1	0.180	00:10	Volume Not Met
Breath Volume		0.171	00:10	
Air Blank		0.000	00:11	
Air Blank		0.000	00:13	
Subject Sample	# 2		00:13	Subject Test Refused
Breath Volume		0.000	00:13	
Air Blank		0.000	00:14	
Control Test		0.080	00:14	
Air Blank		0.000	00:15	
Diagnostics Check		OK	00:15	

Data Download Date/Time: 09/28/2017 13:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDEN,WILLIAM,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: KERN,ETHAN
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 09/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: ETHERIDGE

DATE/TIME OF TEST: 09/26/2017 04:59:10

Diagnostics Check	OK	04:50
Air Blank	0.000	04:50
Control Test	0.081	04:51
Air Blank	0.000	04:51
Subject Sample # 1	0.159	04:54
Breath Volume	1.171	04:54
Air Blank	0.000	04:54
Air Blank	0.000	04:56
Subject Sample # 2	0.166	04:57
Breath Volume	1.203	04:57
Air Blank	0.000	04:58
Control Test	0.080	04:58
Air Blank	0.000	04:59
Diagnostics Check	OK	04:59

Data Download Date/Time: 09/28/2017 13:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDEN,WILLIAM,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HUDSON,CODY
ARRESTING OFFICER AGENCY: Winter Garden PD
LAST AGENCY INSPECTION: 09/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: DOLAN

DATE/TIME OF TEST: 09/27/2017 00:12:58

Diagnostics Check	OK	00:07	
Air Blank	0.000	00:08	
Control Test	0.081	00:08	
Air Blank	0.000	00:09	
Subject Sample	# 1	00:11	Subject Test Refused
Breath Volume	0.000	00:11	
Air Blank	0.000	00:12	
Control Test	0.081	00:12	
Air Blank	0.000	00:12	
Diagnostics Check	OK	00:12	

Data Download Date/Time: 09/28/2017 13:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDEN,WILLIAM,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: AZEVEDO,FABIO
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 09/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:33

SUBJECT'S LAST NAME: RAMPERSAD

DATE/TIME OF TEST: 09/27/2017 23:59:38

Diagnostics Check	OK	23:53
Air Blank	0.000	23:53
Control Test	0.081	23:54
Air Blank	0.000	23:54
Subject Sample # 1	0.187	23:55
Breath Volume	1.359	23:55
Air Blank	0.000	23:55
Air Blank	0.000	23:57
Subject Sample # 2	0.189	23:58
Breath Volume	1.835	23:58
Air Blank	0.000	23:58
Control Test	0.081	23:59
Air Blank	0.000	23:59
Diagnostics Check	OK	23:59

Data Download Date/Time: 09/28/2017 13:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: IRWIN,CHIP
ARRESTING OFFICER AGENCY: Windermere PD
LAST AGENCY INSPECTION: 09/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:22

SUBJECT'S LAST NAME: BOLLING

DATE/TIME OF TEST: 09/30/2017 03:57:05

Diagnostics Check		OK	03:48
Air Blank		0.000	03:49
Control Test		0.081	03:49
Air Blank		0.000	03:50
Subject Sample	# 1	0.198	03:52
Breath Volume		2.429	03:52
Air Blank		0.000	03:52
Air Blank		0.000	03:54
Subject Sample	# 2	0.201	03:55
Breath Volume		2.269	03:55
Air Blank		0.000	03:56
Control Test		0.080	03:56
Air Blank		0.000	03:56
Diagnostics Check		OK	03:57

Data Download Date/Time: 10/05/2017 15:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: AZEVEDO,FABIO
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 09/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:22

SUBJECT'S LAST NAME: RODRIGUEZZAMBRAN

DATE/TIME OF TEST: 09/30/2017 04:53:03

Diagnostics Check	OK	04:46
Air Blank	0.000	04:47
Control Test	0.081	04:47
Air Blank	0.000	04:48
Subject Sample # 1	0.127	04:48
Breath Volume	3.265	04:48
Air Blank	0.000	04:49
Air Blank	0.000	04:51
Subject Sample # 2	0.122	04:51
Breath Volume	2.820	04:51
Air Blank	0.000	04:52
Control Test	0.081	04:52
Air Blank	0.000	04:52
Diagnostics Check	OK	04:53

Data Download Date/Time: 10/05/2017 15:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRAUNLICH,G,A
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: WARREN,C
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 08/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09117080A2
DRY GAS EXPIRY DATE: 06/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:52

SUBJECT'S LAST NAME: THORNE

DATE/TIME OF TEST: 09/05/2017 00:40:18

Diagnostics Check	OK	00:37	
Air Blank	0.000	00:37	
Control Test	0.080	00:37	
Air Blank	0.000	00:38	
Subject Sample	# 1	00:38	Interferent Detect
Breath Volume	1.906	00:38	
Air Blank	0.000	00:39	
Control Test	0.080	00:39	
Air Blank	0.000	00:40	
Diagnostics Check	OK	00:40	

Data Download Date/Time: 09/22/2017 11:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,JOHN,K
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: THOMPSON,JOHN
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 09/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09117080A2
DRY GAS EXPIRY DATE: 06/05/2019
VIOLATION CODE: COURT ORD
ARREST TIME: 09:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:04

SUBJECT'S LAST NAME: NAAS

DATE/TIME OF TEST: 09/28/2017 09:33:14

Diagnostics Check	OK	09:27
Air Blank	0.000	09:27
Control Test	0.080	09:28
Air Blank	0.000	09:28
Subject Sample # 1	0.000	09:29
Breath Volume	2.367	09:29
Air Blank	0.000	09:29
Air Blank	0.000	09:31
Subject Sample # 2	0.000	09:31
Breath Volume	2.328	09:31
Air Blank	0.000	09:32
Control Test	0.080	09:32
Air Blank	0.000	09:33
Diagnostics Check	OK	09:33

Data Download Date/Time: 10/21/2017 10:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ASHE,KEVIN,B
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: NONE,NONE
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 09/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09117080A2
DRY GAS EXPIRY DATE: 06/05/2019
VIOLATION CODE: PROB/PAROL
ARREST TIME: 08:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:30

SUBJECT'S LAST NAME: NAAS

DATE/TIME OF TEST: 09/29/2017 09:04:20

Diagnostics Check	OK	08:58
Air Blank	0.000	08:59
Control Test	0.080	08:59
Air Blank	0.000	08:59
Subject Sample # 1	0.000	09:00
Breath Volume	2.359	09:00
Air Blank	0.000	09:00
Air Blank	0.000	09:02
Subject Sample # 2	0.000	09:02
Breath Volume	2.476	09:02
Air Blank	0.000	09:03
Control Test	0.080	09:03
Air Blank	0.000	09:04
Diagnostics Check	OK	09:04

Data Download Date/Time: 10/21/2017 10:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,EMILY,K
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: FRES,L
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 09/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A2
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: VARADI

DATE/TIME OF TEST: 09/28/2017 00:06:30

Diagnostics Check	OK	00:00
Air Blank	0.000	00:01
Control Test	0.079	00:01
Air Blank	0.000	00:01
Subject Sample # 1	0.000	00:02
Breath Volume	1.464	00:02
Air Blank	0.000	00:02
Air Blank	0.000	00:04
Subject Sample # 2	0.000	00:05
Breath Volume	1.855	00:05
Air Blank	0.000	00:05
Control Test	0.079	00:05
Air Blank	0.000	00:06
Diagnostics Check	OK	00:06

Data Download Date/Time: 10/03/2017 22:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALDO,AARON,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ALLI,ANTHONY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 09/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A2
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: TOMKO

DATE/TIME OF TEST: 09/28/2017 23:17:16

Diagnostics Check	OK	23:11
Air Blank	0.000	23:11
Control Test	0.078	23:11
Air Blank	0.000	23:12
Subject Sample # 1	0.000	23:13
Breath Volume	2.117	23:13
Air Blank	0.000	23:13
Air Blank	0.000	23:15
Subject Sample # 2	0.000	23:15
Breath Volume	1.640	23:15
Air Blank	0.000	23:16
Control Test	0.078	23:16
Air Blank	0.000	23:17
Diagnostics Check	OK	23:17

Data Download Date/Time: 10/03/2017 22:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEISKOPF, MICHAEL, D
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: KARAYIANES, MICHAEL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 09/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A2
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:58

SUBJECT'S LAST NAME: MOTTLE

DATE/TIME OF TEST: 09/30/2017 00:22:28

Diagnostics Check	OK	00:19	
Air Blank	0.000	00:19	
Control Test	0.079	00:20	
Air Blank	0.000	00:20	
Subject Sample	# 1	00:21	Subject Test Refused
Breath Volume	0.000	00:21	
Air Blank	0.000	00:21	
Control Test	0.078	00:21	
Air Blank	0.000	00:22	
Diagnostics Check	OK	00:22	

Data Download Date/Time: 10/03/2017 22:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KARAYIANES,MICHAEL,F
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ALLI,ANTHONY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 09/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A2
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 19:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:08

SUBJECT'S LAST NAME: HUMMEL

DATE/TIME OF TEST: 09/30/2017 20:36:39

Diagnostics Check	OK	20:29
Air Blank	0.000	20:30
Control Test	0.078	20:30
Air Blank	0.000	20:31
Subject Sample # 1	0.327	20:32
Breath Volume	3.714	20:32
Air Blank	0.000	20:32
Air Blank	0.000	20:34
Subject Sample # 2	0.307	20:35
Breath Volume	3.726	20:35
Air Blank	0.000	20:35
Control Test	0.076	20:36
Air Blank	0.000	20:36
Diagnostics Check	OK	20:36

Data Download Date/Time: 10/03/2017 22:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001654

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROIG,DAWN,M
BREATH TEST OPERATOR AGENCY: Miami Shores PD
ARRESTING OFFICER: ROIG,DAWN
ARRESTING OFFICER AGENCY: Miami Shores PD
LAST AGENCY INSPECTION: 09/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:42

SUBJECT'S LAST NAME: FERNANDEZ

DATE/TIME OF TEST: 09/16/2017 02:14:55

Diagnostics Check	OK	02:07
Air Blank	0.000	02:07
Control Test	0.081	02:07
Air Blank	0.000	02:08
Subject Sample # 1	0.237	02:10
Breath Volume	1.621	02:10
Air Blank	0.000	02:10
Air Blank	0.000	02:12
Subject Sample # 2	0.229	02:13
Breath Volume	1.734	02:13
Air Blank	0.000	02:14
Control Test	0.080	02:14
Air Blank	0.000	02:14
Diagnostics Check	OK	02:14

Data Download Date/Time: 10/30/2017 23:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001654

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRITO,KENNHT,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: BRITO,KENNHT
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 09/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:43

SUBJECT'S LAST NAME: MEYER

DATE/TIME OF TEST: 09/16/2017 02:59:03

Diagnostics Check	OK	02:46	
Air Blank	0.000	02:47	
Control Test	0.081	02:47	
Air Blank	0.000	02:47	
Subject Sample	# 1	02:51	No Sample Provided
Breath Volume	0.000	02:51	
Air Blank	0.000	02:51	
Air Blank	0.000	02:53	
Subject Sample	# 2	0.000	02:54
Breath Volume	1.238	02:54	
Air Blank	0.000	02:55	
Air Blank	0.000	02:57	
Subject Sample	# 3	0.000	02:57
Breath Volume	1.453	02:57	
Air Blank	0.000	02:58	
Control Test	0.080	02:58	
Air Blank	0.000	02:58	
Diagnostics Check	OK	02:59	

Data Download Date/Time: 10/30/2017 23:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001654

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BELLINGER, TI'ANDRE, H

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: BELLINGER, TI'ANDRE

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 09/13/2017

Date or Time Changed: No

DRY GAS STANDARD LOT #: 203498

DRY GAS EXPIRY DATE: 12/10/2018

VIOLATION CODE: DUI

ARREST TIME: 04:05

GENDER: Male

OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 09/17/2017 05:05:32

Diagnostics Check	OK	04:49	
Air Blank	0.000	04:50	
Control Test	0.082	04:50	
Air Blank	0.000	04:51	
Subject Sample	# 1 0.179	04:52	
Breath Volume	1.304	04:52	
Air Blank	0.000	04:53	
Air Blank	0.000	04:55	
Subject Sample	# 2	04:58	No Sample Provided
Breath Volume	0.000	04:58	
Air Blank	0.000	04:59	
Air Blank	0.000	05:00	
Subject Sample	# 3	05:03	Subject Test Refused
Breath Volume	0.082	05:03	
Air Blank	0.000	05:04	
Control Test	0.080	05:04	
Air Blank	0.000	05:05	
Diagnostics Check	OK	05:05	

Data Download Date/Time: 10/30/2017 23:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001654

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOPEZ,M,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: LOPEZ,M
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 09/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:44

SUBJECT'S LAST NAME: CONDOR

DATE/TIME OF TEST: 09/29/2017 03:37:51

Diagnostics Check		OK	03:22	
Air Blank		0.000	03:23	
Control Test		0.081	03:23	
Air Blank		0.000	03:24	
Subject Sample	# 1		03:27	No Sample Provided
Breath Volume		0.000	03:27	
Air Blank		0.000	03:27	
Air Blank		0.000	03:29	
Subject Sample	# 2	0.190	03:33	
Breath Volume		2.687	03:33	
Air Blank		0.000	03:33	
Air Blank		0.000	03:35	
Subject Sample	# 3	0.179	03:36	
Breath Volume		2.378	03:36	
Air Blank		0.000	03:36	
Control Test		0.080	03:37	
Air Blank		0.000	03:37	
Diagnostics Check		OK	03:37	

Data Download Date/Time: 10/30/2017 23:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001657

INSTRUMENT REGISTERED TO: MICCOSUKEE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PROZZILLO,ANTHONY,M
BREATH TEST OPERATOR AGENCY: Miccosukee PD
ARRESTING OFFICER: ANTERIO,CHRISTOPHER
ARRESTING OFFICER AGENCY: Miccosukee PD
LAST AGENCY INSPECTION: 08/16/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:30

SUBJECT'S LAST NAME: LAMBERT

DATE/TIME OF TEST: 09/06/2017 21:02:27

Diagnostics Check	OK	20:56
Air Blank	0.000	20:57
Control Test	0.079	20:57
Air Blank	0.000	20:57
Subject Sample # 1	0.000	20:58
Breath Volume	1.421	20:58
Air Blank	0.000	20:58
Air Blank	0.000	21:00
Subject Sample # 2	0.000	21:01
Breath Volume	2.359	21:01
Air Blank	0.000	21:01
Control Test	0.078	21:01
Air Blank	0.000	21:02
Diagnostics Check	OK	21:02

Data Download Date/Time: 10/02/2017 13:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001657

INSTRUMENT REGISTERED TO: MICCOSUKEE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUTIERREZ, YUSNIER,
BREATH TEST OPERATOR AGENCY: Miccosukee PD
ARRESTING OFFICER: RODRIGUEZ, NELSON
ARRESTING OFFICER AGENCY: Miccosukee PD
LAST AGENCY INSPECTION: 08/16/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 06:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:20

SUBJECT'S LAST NAME: QUINONES

DATE/TIME OF TEST: 09/22/2017 07:43:27

Diagnostics Check	OK	07:40	
Air Blank	0.000	07:40	
Control Test	0.078	07:41	
Air Blank	0.000	07:41	
Subject Sample	# 1	07:42	Subject Test Refused
Breath Volume	0.000	07:42	
Air Blank	0.000	07:42	
Control Test	0.079	07:42	
Air Blank	0.000	07:43	
Diagnostics Check	OK	07:43	

Data Download Date/Time: 10/02/2017 13:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001657

INSTRUMENT REGISTERED TO: MICCOSUKEE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUTIERREZ, YUSNIER,
BREATH TEST OPERATOR AGENCY: Miccosukee PD
ARRESTING OFFICER: RODRIGUEZ, NELSON
ARRESTING OFFICER AGENCY: Miccosukee PD
LAST AGENCY INSPECTION: 08/16/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 06:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:20

SUBJECT'S LAST NAME: QUINONES

DATE/TIME OF TEST: 09/22/2017 07:54:11

Diagnostics Check	OK	07:51	
Air Blank	0.000	07:51	
Control Test	0.078	07:52	
Air Blank	0.000	07:52	
Subject Sample	# 1	07:52	Subject Test Refused
Breath Volume	0.000	07:52	
Air Blank	0.000	07:53	
Control Test	0.078	07:53	
Air Blank	0.000	07:54	
Diagnostics Check	OK	07:54	

Data Download Date/Time: 10/02/2017 13:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LYONS,D,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: LYONS,D
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 08/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 19:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:02

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 09/01/2017 19:28:44

Diagnostics Check	OK	19:22
Air Blank	0.000	19:22
Control Test	0.079	19:23
Air Blank	0.000	19:23
Subject Sample # 1	0.065	19:24
Breath Volume	3.089	19:24
Air Blank	0.000	19:24
Air Blank	0.000	19:26
Subject Sample # 2	0.057	19:27
Breath Volume	1.781	19:27
Air Blank	0.000	19:27
Control Test	0.079	19:28
Air Blank	0.000	19:28
Diagnostics Check	OK	19:28

Data Download Date/Time: 09/28/2017 17:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LYONS,D,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: LYONS,D
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 08/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: ADMIN
ARREST TIME: 17:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:48

SUBJECT'S LAST NAME: TORRES

DATE/TIME OF TEST: 09/02/2017 19:14:34

Diagnostics Check	OK	19:08
Air Blank	0.000	19:09
Control Test	0.079	19:09
Air Blank	0.000	19:09
Subject Sample # 1	0.000	19:10
Breath Volume	3.097	19:10
Air Blank	0.000	19:10
Air Blank	0.000	19:12
Subject Sample # 2	0.000	19:13
Breath Volume	3.246	19:13
Air Blank	0.000	19:13
Control Test	0.078	19:13
Air Blank	0.000	19:14
Diagnostics Check	OK	19:14

Data Download Date/Time: 09/28/2017 17:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LYONS,D,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: LYONS,D
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 08/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 04:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:13

SUBJECT'S LAST NAME: VERGARA-TORRES

DATE/TIME OF TEST: 09/03/2017 04:37:14

Diagnostics Check	OK	04:33	
Air Blank	0.000	04:34	
Control Test	0.080	04:34	
Air Blank	0.000	04:34	
Subject Sample	# 1	04:35	Subject Test Refused
Breath Volume	0.000	04:35	
Air Blank	0.000	04:36	
Control Test	0.080	04:36	
Air Blank	0.000	04:37	
Diagnostics Check	OK	04:37	

Data Download Date/Time: 09/28/2017 17:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LYONS,D,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: LYONS,D
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 08/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 19:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:52

SUBJECT'S LAST NAME: IGLESIAS

DATE/TIME OF TEST: 09/03/2017 19:18:28

Diagnostics Check	OK	19:12	
Air Blank	0.000	19:13	
Control Test	0.079	19:13	
Air Blank	0.000	19:14	
Subject Sample # 1	0.189	19:14	
Breath Volume	1.785	19:14	
Air Blank	0.000	19:15	
Air Blank	0.000	19:17	
Subject Sample # 2	0.190	19:17	
Breath Volume	1.796	19:17	
Air Blank		19:18	RFI Detect
Air Blank	0.000	19:18	

Data Download Date/Time: 09/28/2017 17:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LYONS,D,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: LYONS,D
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 08/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 19:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:52

SUBJECT'S LAST NAME: IGLESIAS

DATE/TIME OF TEST: 09/03/2017 19:26:24

Diagnostics Check	OK	19:20
Air Blank	0.000	19:20
Control Test	0.079	19:21
Air Blank	0.000	19:21
Subject Sample # 1	0.189	19:21
Breath Volume	2.054	19:21
Air Blank	0.000	19:22
Air Blank	0.000	19:24
Subject Sample # 2	0.184	19:24
Breath Volume	1.773	19:24
Air Blank	0.000	19:25
Control Test	0.078	19:25
Air Blank	0.000	19:26
Diagnostics Check	OK	19:26

Data Download Date/Time: 09/28/2017 17:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LYONS,D,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: LYONS,D
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 08/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 04:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: ROJAS

DATE/TIME OF TEST: 09/04/2017 04:22:39

Diagnostics Check	OK	04:16
Air Blank	0.000	04:16
Control Test	0.079	04:16
Air Blank	0.000	04:17
Subject Sample # 1	0.174	04:18
Breath Volume	1.839	04:18
Air Blank	0.000	04:18
Air Blank	0.000	04:20
Subject Sample # 2	0.174	04:21
Breath Volume	1.429	04:21
Air Blank	0.000	04:21
Control Test	0.078	04:22
Air Blank	0.000	04:22
Diagnostics Check	OK	04:22

Data Download Date/Time: 09/28/2017 17:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LYONS,D,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: LYONS,D
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 08/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 00:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: MARTIN

DATE/TIME OF TEST: 09/07/2017 01:01:48

Diagnostics Check	OK	00:55
Air Blank	0.000	00:56
Control Test	0.080	00:56
Air Blank	0.000	00:56
Subject Sample # 1	0.074	00:57
Breath Volume	1.234	00:57
Air Blank	0.000	00:57
Air Blank	0.000	00:59
Subject Sample # 2	0.075	01:00
Breath Volume	2.007	01:00
Air Blank	0.000	01:00
Control Test	0.080	01:01
Air Blank	0.000	01:01
Diagnostics Check	OK	01:01

Data Download Date/Time: 09/28/2017 17:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LYONS,D,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: LYONS,D
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 08/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 23:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: ;6360102325588573202

DATE/TIME OF TEST: 09/10/2017 00:11:27

Diagnostics Check	OK	00:05
Air Blank	0.000	00:05
Control Test	0.079	00:06
Air Blank	0.000	00:06
Subject Sample # 1	0.136	00:07
Breath Volume	1.980	00:07
Air Blank	0.000	00:07
Air Blank	0.000	00:09
Subject Sample # 2	0.138	00:09
Breath Volume	2.023	00:09
Air Blank	0.000	00:10
Control Test	0.078	00:10
Air Blank	0.000	00:11
Diagnostics Check	OK	00:11

Data Download Date/Time: 09/28/2017 17:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LYONS,D,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: LYONS,D
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 08/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 01:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: DOLPHUS

DATE/TIME OF TEST: 09/22/2017 02:06:36

Diagnostics Check	OK	02:00
Air Blank	0.000	02:00
Control Test	0.080	02:00
Air Blank	0.000	02:01
Subject Sample # 1	0.177	02:02
Breath Volume	3.031	02:02
Air Blank	0.000	02:02
Air Blank	0.000	02:04
Subject Sample # 2	0.179	02:05
Breath Volume	2.945	02:05
Air Blank	0.000	02:05
Control Test	0.078	02:05
Air Blank	0.000	02:06
Diagnostics Check	OK	02:06

Data Download Date/Time: 09/28/2017 17:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001720

INSTRUMENT REGISTERED TO: ALACHUA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BENNETT,JONATHON,K
BREATH TEST OPERATOR AGENCY: Alachua PD
ARRESTING OFFICER: SALTER,ROBERT
ARRESTING OFFICER AGENCY: Alachua PD
LAST AGENCY INSPECTION: 08/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A6
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 13:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 12:25

SUBJECT'S LAST NAME: BIALACH

DATE/TIME OF TEST: 09/02/2017 13:30:40

Diagnostics Check	OK	13:28	
Air Blank	0.000	13:28	
Control Test	0.080	13:28	
Air Blank	0.000	13:29	
Subject Sample	# 1	13:30	Sequence Aborted
Breath Volume	0.000	13:30	
Air Blank	0.000	13:30	

Data Download Date/Time: 10/02/2017 20:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001720

INSTRUMENT REGISTERED TO: ALACHUA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BENNETT,JONATHON,K
BREATH TEST OPERATOR AGENCY: Alachua PD
ARRESTING OFFICER: SALTER,ROBERT
ARRESTING OFFICER AGENCY: Alachua PD
LAST AGENCY INSPECTION: 08/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A6
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 13:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:25

SUBJECT'S LAST NAME: BIALACH

DATE/TIME OF TEST: 09/02/2017 13:52:29

Diagnostics Check	OK	13:45
Air Blank	0.000	13:46
Control Test	0.080	13:46
Air Blank	0.000	13:47
Subject Sample # 1	0.153	13:47
Breath Volume	1.753	13:47
Air Blank	0.000	13:48
Air Blank	0.000	13:50
Subject Sample # 2	0.155	13:50
Breath Volume	2.425	13:50
Air Blank	0.000	13:51
Control Test	0.079	13:51
Air Blank	0.000	13:52
Diagnostics Check	OK	13:52

Data Download Date/Time: 10/02/2017 20:27