

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VISMAN,CLAYTON,A
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: VISMAN,CLAYTON
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 09/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 05:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:37

SUBJECT'S LAST NAME: CANIFF

DATE/TIME OF TEST: 10/01/2017 06:06:37

Diagnostics Check	OK	06:00
Air Blank	0.000	06:00
Control Test	0.079	06:01
Air Blank	0.000	06:01
Subject Sample # 1	0.090	06:02
Breath Volume	2.687	06:02
Air Blank	0.000	06:02
Air Blank	0.000	06:04
Subject Sample # 2	0.086	06:05
Breath Volume	1.988	06:05
Air Blank	0.000	06:05
Control Test	0.079	06:06
Air Blank	0.000	06:06
Diagnostics Check	OK	06:06

Data Download Date/Time: 10/26/2017 09:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MURPHY,JAVON,D
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: MURPHY,JAVON
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 09/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 05:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:47

SUBJECT'S LAST NAME: HALL

DATE/TIME OF TEST: 10/01/2017 06:28:40

Diagnostics Check	OK	06:17	
Air Blank	0.000	06:17	
Control Test	0.080	06:18	
Air Blank	0.000	06:18	
Subject Sample	# 1	06:21	No Sample Provided
Breath Volume	0.000	06:21	
Air Blank	0.000	06:22	
Air Blank	0.000	06:24	
Subject Sample	# 2	06:27	No Sample Provided
Breath Volume	0.000	06:27	
Air Blank	0.000	06:27	
Control Test	0.080	06:28	
Air Blank	0.000	06:28	
Diagnostics Check	OK	06:28	

Data Download Date/Time: 10/26/2017 09:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PINKSTON, RONALD, A
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: SHUTTERLY, BROOKE
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 09/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 14:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:30

SUBJECT'S LAST NAME: DOUGHERTY

DATE/TIME OF TEST: 10/06/2017 15:52:27

Diagnostics Check	OK	15:52	
Air Blank		15:52	RFI Detect
Air Blank		15:52	RFI Detect

Data Download Date/Time: 10/26/2017 09:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PINSKTON, RONALD, A
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: SHUTTERLY, BROOKE
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 09/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 14:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:30

SUBJECT'S LAST NAME: DOUGHERTY

DATE/TIME OF TEST: 10/06/2017 16:07:49

Diagnostics Check	OK	15:56	
Air Blank	0.000	15:56	
Control Test	0.080	15:57	
Air Blank	0.000	15:57	
Subject Sample	# 1	16:00	No Sample Provided
Breath Volume	0.000	16:00	
Air Blank	0.000	16:01	
Air Blank	0.000	16:03	
Subject Sample	# 2	16:06	No Sample Provided
Breath Volume	0.000	16:06	
Air Blank	0.000	16:06	
Control Test	0.080	16:07	
Air Blank	0.000	16:07	
Diagnostics Check	OK	16:07	

Data Download Date/Time: 10/26/2017 09:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWARD,ANTHONY,M
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: BROWM,MYKAHEL
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 09/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: BUSCH

DATE/TIME OF TEST: 10/08/2017 02:41:16

Diagnostics Check	OK	02:34
Air Blank	0.000	02:35
Control Test	0.080	02:35
Air Blank	0.000	02:36
Subject Sample # 1	0.119	02:36
Breath Volume	3.546	02:36
Air Blank	0.000	02:37
Air Blank	0.000	02:39
Subject Sample # 2	0.122	02:39
Breath Volume	3.089	02:39
Air Blank	0.000	02:40
Control Test	0.078	02:40
Air Blank	0.000	02:41
Diagnostics Check	OK	02:41

Data Download Date/Time: 10/26/2017 09:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNS,WILLIAM,A
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: MORRIS-BROWN,MYKHEL
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 09/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:17

SUBJECT'S LAST NAME: GORDON

DATE/TIME OF TEST: 10/08/2017 23:51:56

Diagnostics Check	OK	23:44
Air Blank	0.000	23:44
Control Test	0.081	23:45
Air Blank	0.000	23:45
Subject Sample # 1	0.195	23:47
Breath Volume	1.417	23:47
Air Blank	0.000	23:48
Air Blank	0.000	23:49
Subject Sample # 2	0.190	23:50
Breath Volume	1.468	23:50
Air Blank	0.000	23:51
Control Test	0.076	23:51
Air Blank	0.000	23:51
Diagnostics Check	OK	23:51

Data Download Date/Time: 10/26/2017 09:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWARD,ANTHONY,M
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: HUGHES,DAVID
ARRESTING OFFICER AGENCY: Alachua PD
LAST AGENCY INSPECTION: 09/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:39

SUBJECT'S LAST NAME: WALLACE

DATE/TIME OF TEST: 10/09/2017 05:13:25

Diagnostics Check	OK	05:09	
Air Blank	0.000	05:09	
Control Test	0.080	05:10	
Air Blank	0.000	05:10	
Subject Sample	# 1	05:12	Subject Test Refused
Breath Volume	0.000	05:12	
Air Blank	0.000	05:12	
Control Test	0.081	05:12	
Air Blank	0.000	05:13	
Diagnostics Check	OK	05:13	

Data Download Date/Time: 10/26/2017 09:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VISMAN,CLAYTON,A
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: VISMAN,CLAYTON
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 09/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 11:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: PAYTON

DATE/TIME OF TEST: 10/10/2017 01:00:07

Diagnostics Check	OK	00:53
Air Blank	0.000	00:54
Control Test	0.079	00:54
Air Blank	0.000	00:54
Subject Sample # 1	0.186	00:55
Breath Volume	2.628	00:55
Air Blank	0.000	00:56
Air Blank	0.000	00:57
Subject Sample # 2	0.170	00:58
Breath Volume	2.648	00:58
Air Blank	0.000	00:59
Control Test	0.080	00:59
Air Blank	0.000	01:00
Diagnostics Check	OK	01:00

Data Download Date/Time: 10/26/2017 09:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNS,WILLIAM,A
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: WILLIAMS,JUSTIN
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 09/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: GOFF

DATE/TIME OF TEST: 10/12/2017 04:28:16

Diagnostics Check	OK	04:21
Air Blank	0.000	04:22
Control Test	0.080	04:22
Air Blank	0.000	04:23
Subject Sample # 1	0.102	04:23
Breath Volume	5.417	04:23
Air Blank	0.000	04:24
Air Blank	0.000	04:26
Subject Sample # 2	0.102	04:26
Breath Volume	5.660	04:26
Air Blank	0.000	04:27
Control Test	0.078	04:27
Air Blank	0.000	04:28
Diagnostics Check	OK	04:28

Data Download Date/Time: 10/26/2017 09:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VISMAN,CLAYTON,A
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: SMITH,CRAIG
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 09/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: GORMLEY

DATE/TIME OF TEST: 10/14/2017 22:29:58

Diagnostics Check	OK	22:27	
Air Blank	0.000	22:27	
Control Test	0.079	22:27	
Air Blank	0.000	22:28	
Subject Sample	# 1	22:28	Subject Test Refused
Breath Volume	0.000	22:28	
Air Blank	0.000	22:29	
Control Test	0.080	22:29	
Air Blank	0.000	22:29	
Diagnostics Check	OK	22:29	

Data Download Date/Time: 10/26/2017 09:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNS,WILLIAM,A
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: ADKINS,MICHAEL
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 09/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:43

SUBJECT'S LAST NAME: HADAWAY

DATE/TIME OF TEST: 10/19/2017 06:15:05

Diagnostics Check	OK	06:05
Air Blank	0.000	06:06
Control Test	0.081	06:06
Air Blank	0.000	06:07
Subject Sample # 1	0.205	06:07
Breath Volume	1.886	06:07
Air Blank	0.000	06:08
Air Blank	0.000	06:10
Subject Sample # 2	0.202	06:13
Breath Volume	1.421	06:13
Air Blank	0.000	06:14
Control Test	0.078	06:14
Air Blank	0.000	06:15
Diagnostics Check	OK	06:15

Data Download Date/Time: 10/26/2017 09:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUNT,CHARLES,R
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: HUNT,CHARLES
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 09/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:28

SUBJECT'S LAST NAME: GLYNN

DATE/TIME OF TEST: 10/21/2017 04:05:03

Diagnostics Check	OK	03:55	
Air Blank	0.000	03:55	
Control Test	0.079	03:56	
Air Blank	0.000	03:56	
Subject Sample	# 1	03:59	No Sample Provided
Breath Volume	0.000	03:59	
Air Blank	0.000	04:00	
Air Blank	0.000	04:02	
Subject Sample	# 2	04:03	Subject Test Refused
Breath Volume	0.000	04:03	
Air Blank	0.000	04:04	
Control Test	0.079	04:04	
Air Blank	0.000	04:04	
Diagnostics Check	OK	04:05	

Data Download Date/Time: 10/26/2017 09:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWARD,ANTHONY,M
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: ROCA,DANIEL
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 09/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: KESSELRING

DATE/TIME OF TEST: 10/22/2017 03:03:54

Diagnostics Check	OK	02:57
Air Blank	0.000	02:57
Control Test	0.080	02:57
Air Blank	0.000	02:58
Subject Sample # 1	0.194	02:59
Breath Volume	2.316	02:59
Air Blank	0.000	02:59
Air Blank	0.000	03:01
Subject Sample # 2	0.201	03:02
Breath Volume	2.074	03:02
Air Blank	0.000	03:02
Control Test	0.078	03:03
Air Blank	0.000	03:03
Diagnostics Check	OK	03:03

Data Download Date/Time: 10/26/2017 09:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICHARDSON, MICHAEL, R
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: BEVILLE, TRAVIS
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 10/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:42

SUBJECT'S LAST NAME: SCOTT

DATE/TIME OF TEST: 10/29/2017 03:09:06

Diagnostics Check	OK	03:03
Air Blank	0.000	03:03
Control Test	0.080	03:03
Air Blank	0.000	03:04
Subject Sample # 1	0.130	03:04
Breath Volume	1.359	03:04
Air Blank	0.000	03:05
Air Blank	0.000	03:07
Subject Sample # 2	0.129	03:07
Breath Volume	1.941	03:07
Air Blank	0.000	03:08
Control Test	0.078	03:08
Air Blank	0.000	03:09
Diagnostics Check	OK	03:09

Data Download Date/Time: 11/28/2017 21:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VISMAN,CLAYTON,A
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: VISMAN,CLAYTON
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 10/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:15

SUBJECT'S LAST NAME: KAYLOR

DATE/TIME OF TEST: 10/29/2017 21:47:26

Diagnostics Check	OK	21:38	
Air Blank	0.000	21:38	
Control Test	0.080	21:39	
Air Blank	0.000	21:39	
Subject Sample	# 1	21:42	No Sample Provided
Breath Volume	0.000	21:42	
Air Blank	0.000	21:43	
Air Blank	0.000	21:45	
Subject Sample	# 2	21:46	Subject Test Refused
Breath Volume	0.000	21:46	
Air Blank	0.000	21:46	
Control Test	0.081	21:46	
Air Blank	0.000	21:47	
Diagnostics Check	OK	21:47	

Data Download Date/Time: 11/28/2017 21:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOLL, MICHAEL,
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: WOLL, MICHAEL
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 10/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 22:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:54

SUBJECT'S LAST NAME: GEIGER

DATE/TIME OF TEST: 10/27/2017 00:26:48

Diagnostics Check	OK	00:20	
Air Blank	0.000	00:21	
Control Test	0.081	00:21	
Air Blank	0.000	00:22	
Subject Sample # 1	0.258	00:22	
Breath Volume	2.609	00:22	
Air Blank	0.000	00:23	
Air Blank	0.000	00:25	
Subject Sample # 2		00:25	Subject Test Refused
Breath Volume	0.000	00:25	
Air Blank	0.000	00:25	
Control Test	0.079	00:26	
Air Blank	0.000	00:26	
Diagnostics Check	OK	00:26	

Data Download Date/Time: 11/28/2017 16:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURNHAM,J,D
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: BURNHAM,J
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 10/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 02:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:04

SUBJECT'S LAST NAME: PEREA-MARTINEZ

DATE/TIME OF TEST: 10/27/2017 03:39:41

Diagnostics Check	OK	03:26	
Air Blank	0.000	03:27	
Control Test	0.080	03:27	
Air Blank	0.000	03:28	
Subject Sample	# 1 0.232	03:29	
Breath Volume	1.281	03:29	
Air Blank	0.000	03:29	
Air Blank	0.000	03:31	
Subject Sample	# 2 0.156	03:34	Volume Not Met
Breath Volume	0.074	03:34	
Air Blank	0.000	03:35	
Air Blank	0.000	03:37	
Subject Sample	# 3 0.214	03:38	
Breath Volume	2.164	03:38	
Air Blank	0.000	03:38	
Control Test	0.078	03:39	
Air Blank	0.000	03:39	
Diagnostics Check	OK	03:39	

Data Download Date/Time: 11/28/2017 16:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPEAKMAN,ROBERT,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: SPEAKMAN,ROBERT
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 10/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 05:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:06

SUBJECT'S LAST NAME: SHEHEE

DATE/TIME OF TEST: 10/27/2017 06:32:04

Diagnostics Check	OK	06:29	
Air Blank	0.000	06:29	
Control Test	0.079	06:29	
Air Blank	0.000	06:30	
Subject Sample	# 1	06:30	Subject Test Refused
Breath Volume	0.000	06:30	
Air Blank	0.000	06:31	
Control Test	0.080	06:31	
Air Blank	0.000	06:31	
Diagnostics Check	OK	06:32	

Data Download Date/Time: 11/28/2017 16:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILHELM,TONY,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: DRANE,RICHARD
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 10/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 04:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:45

SUBJECT'S LAST NAME: LIPSCOMB

DATE/TIME OF TEST: 10/28/2017 05:08:29

Diagnostics Check	OK	05:05	
Air Blank	0.000	05:06	
Control Test	0.080	05:06	
Air Blank	0.000	05:06	
Subject Sample	# 1	05:07	Subject Test Refused
Breath Volume	0.000	05:07	
Air Blank	0.000	05:07	
Control Test	0.080	05:07	
Air Blank	0.000	05:08	
Diagnostics Check	OK	05:08	

Data Download Date/Time: 11/28/2017 16:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: USRY,STEPHEN,K
BREATH TEST OPERATOR AGENCY: Niceville PD
ARRESTING OFFICER: USRY,STEPHEN
ARRESTING OFFICER AGENCY: Niceville PD
LAST AGENCY INSPECTION: 10/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 01:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:53

SUBJECT'S LAST NAME: TATUM

DATE/TIME OF TEST: 10/29/2017 03:35:10

Diagnostics Check	OK	03:28
Air Blank	0.000	03:29
Control Test	0.081	03:29
Air Blank	0.000	03:30
Subject Sample # 1	0.199	03:30
Breath Volume	1.281	03:30
Air Blank	0.000	03:31
Air Blank	0.000	03:33
Subject Sample # 2	0.181	03:33
Breath Volume	1.718	03:33
Air Blank	0.000	03:34
Control Test	0.080	03:34
Air Blank	0.000	03:35
Diagnostics Check	OK	03:35

Data Download Date/Time: 11/28/2017 16:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMSON, TYLER, H
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: DRANE, RICHARD
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 10/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 03:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: LINARD

DATE/TIME OF TEST: 10/29/2017 04:22:48

Diagnostics Check	OK	04:16
Air Blank	0.000	04:16
Control Test	0.082	04:16
Air Blank	0.000	04:17
Subject Sample # 1	0.157	04:18
Breath Volume	3.152	04:18
Air Blank	0.000	04:18
Air Blank	0.000	04:20
Subject Sample # 2	0.155	04:21
Breath Volume	3.753	04:21
Air Blank	0.000	04:21
Control Test	0.081	04:22
Air Blank	0.000	04:22
Diagnostics Check	OK	04:22

Data Download Date/Time: 11/28/2017 16:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPEAKMAN,ROBERT,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: SPEAKMAN,ROBERT
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 10/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 08:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:57

SUBJECT'S LAST NAME: DAY

DATE/TIME OF TEST: 10/29/2017 09:33:22

Diagnostics Check	OK	09:30	
Air Blank	0.000	09:30	
Control Test	0.081	09:31	
Air Blank	0.000	09:31	
Subject Sample	# 1	09:31	Subject Test Refused
Breath Volume	0.000	09:31	
Air Blank	0.000	09:32	
Control Test	0.081	09:32	
Air Blank	0.000	09:33	
Diagnostics Check	OK	09:33	

Data Download Date/Time: 11/28/2017 16:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILHELM,TONY,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: DRANE,RICHARD
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 10/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 21:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: MERCADO

DATE/TIME OF TEST: 10/29/2017 22:04:29

Diagnostics Check	OK	22:01	
Air Blank	0.000	22:02	
Control Test	0.082	22:02	
Air Blank	0.000	22:02	
Subject Sample	# 1	22:03	Subject Test Refused
Breath Volume	0.000	22:03	
Air Blank	0.000	22:03	
Control Test	0.082	22:03	
Air Blank	0.000	22:04	
Diagnostics Check	OK	22:04	

Data Download Date/Time: 11/28/2017 16:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILHELM,TONY,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: FEENEY,ERIC
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 10/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 03:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:25

SUBJECT'S LAST NAME: SPEARS

DATE/TIME OF TEST: 10/30/2017 04:52:12

Diagnostics Check	OK	04:49	
Air Blank	0.000	04:49	
Control Test	0.081	04:50	
Air Blank	0.000	04:50	
Subject Sample	# 1	04:50	Subject Test Refused
Breath Volume	0.000	04:50	
Air Blank	0.000	04:51	
Control Test	0.081	04:51	
Air Blank	0.000	04:52	
Diagnostics Check	OK	04:52	

Data Download Date/Time: 11/28/2017 16:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000742

INSTRUMENT REGISTERED TO: VALPARAISO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAYLOR,RICHARD,L
BREATH TEST OPERATOR AGENCY: Valparaiso PD
ARRESTING OFFICER: HOUGH,BRYAN
ARRESTING OFFICER AGENCY: Valparaiso PD
LAST AGENCY INSPECTION: 09/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 770784
DRY GAS EXPIRY DATE: 04/27/2020
VIOLATION CODE: DUI
ARREST TIME: 00:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:12

SUBJECT'S LAST NAME: CAMPBELL

DATE/TIME OF TEST: 10/21/2017 01:45:11

Diagnostics Check	OK	01:36	
Air Blank	0.000	01:37	
Control Test	0.080	01:37	
Air Blank	0.000	01:38	
Subject Sample	# 1	01:41	No Sample Provided
Breath Volume	0.000	01:41	
Air Blank	0.000	01:41	
Air Blank	0.000	01:43	
Subject Sample	# 2	01:43	Subject Test Refused
Breath Volume	0.000	01:43	
Air Blank	0.000	01:44	
Control Test	0.081	01:44	
Air Blank	0.000	01:45	
Diagnostics Check	OK	01:45	

Data Download Date/Time: 10/27/2017 14:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALVAREZ,ERIC,C
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: SECOR,BROC
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 09/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 10/01/2017 03:05:58

Diagnostics Check	OK	03:02	
Air Blank	0.000	03:03	
Control Test	0.077	03:03	
Air Blank	0.000	03:04	
Subject Sample	# 1	03:04	Subject Test Refused
Breath Volume	0.000	03:04	
Air Blank	0.000	03:05	
Control Test	0.077	03:05	
Air Blank	0.000	03:05	
Diagnostics Check	OK	03:05	

Data Download Date/Time: 10/16/2017 09:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENLEY, WILLIAM, C
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: HENLEY, WILLIAM
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 09/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:23

SUBJECT'S LAST NAME: ANDERSEN

DATE/TIME OF TEST: 10/04/2017 21:52:15

Diagnostics Check	OK	21:43
Air Blank	0.000	21:43
Control Test	0.079	21:44
Air Blank	0.000	21:44
Subject Sample # 1	0.114	21:47
Breath Volume	2.027	21:47
Air Blank	0.000	21:47
Air Blank	0.000	21:49
Subject Sample # 2	0.114	21:50
Breath Volume	2.269	21:50
Air Blank	0.000	21:51
Control Test	0.077	21:51
Air Blank	0.000	21:52
Diagnostics Check	OK	21:52

Data Download Date/Time: 10/16/2017 09:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHELLEY,SAMUEL,L
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: HAMPTON,LOGAN
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 09/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: GIPSON

DATE/TIME OF TEST: 10/09/2017 22:37:34

Diagnostics Check	OK	22:30
Air Blank	0.000	22:31
Control Test	0.078	22:31
Air Blank	0.000	22:31
Subject Sample # 1	0.154	22:32
Breath Volume	3.515	22:32
Air Blank	0.000	22:33
Air Blank	0.000	22:35
Subject Sample # 2	0.154	22:36
Breath Volume	3.457	22:36
Air Blank	0.000	22:36
Control Test	0.078	22:36
Air Blank	0.000	22:37
Diagnostics Check	OK	22:37

Data Download Date/Time: 10/16/2017 09:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHELLEY,SAMUEL,L
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: SHELLEY,SAMUEL
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 09/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: AL-JAZRA

DATE/TIME OF TEST: 10/10/2017 01:12:40

Diagnostics Check	OK	01:06
Air Blank	0.000	01:06
Control Test	0.077	01:06
Air Blank	0.000	01:07
Subject Sample # 1	0.110	01:08
Breath Volume	4.675	01:08
Air Blank	0.000	01:08
Air Blank	0.000	01:10
Subject Sample # 2	0.106	01:11
Breath Volume	3.640	01:11
Air Blank	0.000	01:11
Control Test	0.078	01:12
Air Blank	0.000	01:12
Diagnostics Check	OK	01:12

Data Download Date/Time: 10/16/2017 09:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KELLY,PATRICK,D
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: EVANGELISTA,CURTIS
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 09/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: PACYGA

DATE/TIME OF TEST: 10/10/2017 21:57:41

Diagnostics Check	OK	21:51
Air Blank	0.000	21:51
Control Test	0.079	21:52
Air Blank	0.000	21:52
Subject Sample # 1	0.000	21:53
Breath Volume	1.609	21:53
Air Blank	0.000	21:53
Air Blank	0.000	21:55
Subject Sample # 2	0.000	21:56
Breath Volume	2.554	21:56
Air Blank	0.000	21:56
Control Test	0.079	21:57
Air Blank	0.000	21:57
Diagnostics Check	OK	21:57

Data Download Date/Time: 10/16/2017 09:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DILLASHAW,HUNTER,C
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: DILLASHAW,HUNTER
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 09/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:07

SUBJECT'S LAST NAME: ROWE

DATE/TIME OF TEST: 10/14/2017 02:36:06

Diagnostics Check	OK	02:34	
Air Blank	0.000	02:35	
Control Test	0.079	02:35	
Air Blank		02:35	RFI Detect
Air Blank	0.000	02:36	

Data Download Date/Time: 10/16/2017 09:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DILLASHAW,HUNTER,C
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: DILLASHAW,HUNTER
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 09/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:07

SUBJECT'S LAST NAME: ROWE

DATE/TIME OF TEST: 10/14/2017 02:44:44

Diagnostics Check	OK	02:38
Air Blank	0.000	02:38
Control Test	0.077	02:39
Air Blank	0.000	02:39
Subject Sample # 1	0.159	02:40
Breath Volume	2.164	02:40
Air Blank	0.000	02:40
Air Blank	0.000	02:42
Subject Sample # 2	0.148	02:43
Breath Volume	2.363	02:43
Air Blank	0.000	02:43
Control Test	0.077	02:44
Air Blank	0.000	02:44
Diagnostics Check	OK	02:44

Data Download Date/Time: 10/16/2017 09:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VO,SEBASTIAN,D
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: VO,SEBASTIAN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 09/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:21

SUBJECT'S LAST NAME: ENRIQUE

DATE/TIME OF TEST: 10/15/2017 23:03:17

Diagnostics Check	OK	23:02	
Air Blank		23:02	Ambient Fail
Air Blank	0.000	23:03	

Data Download Date/Time: 10/16/2017 09:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VO,SEBASTIAN,D
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: VO,SEBASTIAN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 09/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:21

SUBJECT'S LAST NAME: ENRIQUE

DATE/TIME OF TEST: 10/15/2017 23:05:56

Diagnostics Check	OK	23:04	
Air Blank	0.000	23:05	
Control Test	0.070	23:05	Control Outside Tolerance
Air Blank	0.000	23:05	

Data Download Date/Time: 10/16/2017 09:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHELLEY,SAMUEL,L
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: OSBORNE,JOSH
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 10/16/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: GAUGHAN

DATE/TIME OF TEST: 10/16/2017 22:33:52

Diagnostics Check	OK	22:27
Air Blank	0.000	22:27
Control Test	0.079	22:28
Air Blank	0.000	22:28
Subject Sample # 1	0.201	22:29
Breath Volume	1.332	22:29
Air Blank	0.000	22:29
Air Blank	0.000	22:31
Subject Sample # 2	0.206	22:32
Breath Volume	3.960	22:32
Air Blank	0.000	22:32
Control Test	0.078	22:33
Air Blank	0.000	22:33
Diagnostics Check	OK	22:33

Data Download Date/Time: 11/29/2017 09:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BYRD,BRANDON,C
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: BOATWRIGHT,JAMIE
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 10/16/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:23

SUBJECT'S LAST NAME: DICIANNI

DATE/TIME OF TEST: 10/16/2017 23:59:10

Diagnostics Check	OK	23:52
Air Blank	0.000	23:52
Control Test	0.078	23:53
Air Blank	0.000	23:53
Subject Sample # 1	0.042	23:54
Breath Volume	2.832	23:54
Air Blank	0.000	23:55
Air Blank	0.000	23:57
Subject Sample # 2	0.040	23:57
Breath Volume	2.566	23:57
Air Blank	0.000	23:58
Control Test	0.077	23:58
Air Blank	0.000	23:59
Diagnostics Check	OK	23:59

Data Download Date/Time: 11/29/2017 09:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DILLASHAW,HUNTER,C
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: EVANGELISTA,CURTIS
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 10/16/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:25

SUBJECT'S LAST NAME: MITCHELL

DATE/TIME OF TEST: 10/19/2017 22:52:00

Diagnostics Check	OK	22:45
Air Blank	0.000	22:46
Control Test	0.079	22:46
Air Blank	0.000	22:47
Subject Sample # 1	0.158	22:47
Breath Volume	2.398	22:47
Air Blank	0.000	22:48
Air Blank	0.000	22:50
Subject Sample # 2	0.153	22:50
Breath Volume	1.925	22:50
Air Blank	0.000	22:51
Control Test	0.077	22:51
Air Blank	0.000	22:51
Diagnostics Check	OK	22:52

Data Download Date/Time: 11/29/2017 09:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENLEY, WILLIAM, C
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: HENLEY, WILLIAM
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 10/16/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:05

SUBJECT'S LAST NAME: SALTER

DATE/TIME OF TEST: 10/20/2017 21:35:34

Diagnostics Check	OK	21:25	
Air Blank	0.000	21:25	
Control Test	0.079	21:26	
Air Blank	0.000	21:26	
Subject Sample	# 1 0.139	21:29	Volume Not Met
Breath Volume	0.562	21:29	
Air Blank	0.000	21:30	
Air Blank	0.000	21:32	
Subject Sample	# 2	21:34	Subject Test Refused
Breath Volume	0.316	21:34	
Air Blank	0.000	21:34	
Control Test	0.078	21:34	
Air Blank	0.000	21:35	
Diagnostics Check	OK	21:35	

Data Download Date/Time: 11/29/2017 09:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,PATRICK,A
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: WILLIAMS,PATRICK
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 10/16/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:17

SUBJECT'S LAST NAME: TOMPKINS

DATE/TIME OF TEST: 10/20/2017 23:44:23

Diagnostics Check	OK	23:38
Air Blank	0.000	23:38
Control Test	0.078	23:39
Air Blank	0.000	23:39
Subject Sample # 1	0.088	23:40
Breath Volume	4.460	23:40
Air Blank	0.000	23:40
Air Blank	0.000	23:42
Subject Sample # 2	0.085	23:42
Breath Volume	4.238	23:42
Air Blank	0.000	23:43
Control Test	0.078	23:43
Air Blank	0.000	23:44
Diagnostics Check	OK	23:44

Data Download Date/Time: 11/29/2017 09:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENLEY, WILLIAM, C
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: HENLEY, WILLIAM
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 10/16/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:04

SUBJECT'S LAST NAME: MORRIS

DATE/TIME OF TEST: 10/22/2017 04:33:35

Diagnostics Check	OK	04:28	
Air Blank	0.000	04:28	
Control Test	0.079	04:29	
Air Blank	0.000	04:29	
Subject Sample	# 1	04:32	Subject Test Refused
Breath Volume	0.000	04:32	
Air Blank	0.000	04:32	
Control Test	0.078	04:33	
Air Blank	0.000	04:33	
Diagnostics Check	OK	04:33	

Data Download Date/Time: 11/29/2017 09:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCPHERSON,DAISY,L
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: BOATWRIGHT,JAMIE
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 10/16/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:30

SUBJECT'S LAST NAME: MCPHERSON

DATE/TIME OF TEST: 10/22/2017 20:54:06

Diagnostics Check	OK	20:50	
Air Blank	0.000	20:51	
Control Test	0.078	20:51	
Air Blank	0.000	20:52	
Subject Sample	# 1	20:52	Subject Test Refused
Breath Volume	0.000	20:52	
Air Blank	0.000	20:53	
Control Test	0.077	20:53	
Air Blank	0.000	20:54	
Diagnostics Check	OK	20:54	

Data Download Date/Time: 11/29/2017 09:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHORETTE,BRIAN,E
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: SHORETTE,BRIAN
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 10/16/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:23

SUBJECT'S LAST NAME: WALTER

DATE/TIME OF TEST: 10/25/2017 23:49:37

Diagnostics Check	OK	23:43
Air Blank	0.000	23:43
Control Test	0.079	23:44
Air Blank	0.000	23:44
Subject Sample # 1	0.109	23:45
Breath Volume	1.640	23:45
Air Blank	0.000	23:45
Air Blank	0.000	23:47
Subject Sample # 2	0.117	23:48
Breath Volume	1.617	23:48
Air Blank	0.000	23:48
Control Test	0.077	23:49
Air Blank	0.000	23:49
Diagnostics Check	OK	23:49

Data Download Date/Time: 11/29/2017 09:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARY,ADAM,C
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: DAVIS,AMANDA
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 10/16/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:07

SUBJECT'S LAST NAME: VANDENBERG

DATE/TIME OF TEST: 10/26/2017 02:35:50

Diagnostics Check	OK	02:29
Air Blank	0.000	02:29
Control Test	0.078	02:29
Air Blank	0.000	02:30
Subject Sample # 1	0.192	02:31
Breath Volume	2.464	02:31
Air Blank	0.000	02:32
Air Blank	0.000	02:33
Subject Sample # 2	0.174	02:34
Breath Volume	1.304	02:34
Air Blank	0.000	02:34
Control Test	0.077	02:35
Air Blank	0.000	02:35
Diagnostics Check	OK	02:35

Data Download Date/Time: 11/29/2017 09:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SALON,DENNIS,M
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: SALON,DENNIS
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 10/16/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 16:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:42

SUBJECT'S LAST NAME: DODSON

DATE/TIME OF TEST: 10/27/2017 17:08:58

Diagnostics Check	OK	17:02
Air Blank	0.000	17:03
Control Test	0.077	17:03
Air Blank	0.000	17:04
Subject Sample # 1	0.037	17:04
Breath Volume	2.894	17:04
Air Blank	0.000	17:05
Air Blank	0.000	17:07
Subject Sample # 2	0.035	17:07
Breath Volume	2.531	17:07
Air Blank	0.000	17:08
Control Test	0.077	17:08
Air Blank	0.000	17:08
Diagnostics Check	OK	17:08

Data Download Date/Time: 11/29/2017 09:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHORETTE,BRIAN,E
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: LEITE,STEVEN
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 10/16/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:55

SUBJECT'S LAST NAME: CHOICE

DATE/TIME OF TEST: 10/28/2017 22:21:56

Diagnostics Check	OK	22:16
Air Blank	0.000	22:16
Control Test	0.077	22:16
Air Blank	0.000	22:17
Subject Sample # 1	0.121	22:17
Breath Volume	1.910	22:17
Air Blank	0.000	22:18
Air Blank	0.000	22:20
Subject Sample # 2	0.122	22:20
Breath Volume	1.898	22:20
Air Blank	0.000	22:21
Control Test	0.077	22:21
Air Blank	0.000	22:21
Diagnostics Check	OK	22:21

Data Download Date/Time: 11/29/2017 09:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BYRD,BRANDON,C
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: AIKENS,NATHANIEL
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 10/16/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: BRIMMER

DATE/TIME OF TEST: 10/31/2017 21:45:00

Diagnostics Check	OK	21:37
Air Blank	0.000	21:38
Control Test	0.078	21:38
Air Blank	0.000	21:39
Subject Sample # 1	0.103	21:40
Breath Volume	3.222	21:40
Air Blank	0.000	21:40
Air Blank	0.000	21:42
Subject Sample # 2	0.090	21:43
Breath Volume	1.878	21:43
Air Blank	0.000	21:44
Control Test	0.079	21:44
Air Blank	0.000	21:44
Diagnostics Check	OK	21:45

Data Download Date/Time: 11/29/2017 09:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POSEY,CHETEMA,M
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: BOATWRIGHT,JAMIE
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 10/16/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:54

SUBJECT'S LAST NAME: POSEY

DATE/TIME OF TEST: 10/31/2017 23:25:07

Diagnostics Check	OK	23:14
Air Blank	0.000	23:15
Control Test	0.078	23:15
Air Blank	0.000	23:16
Subject Sample # 1	0.236	23:16
Breath Volume	1.714	23:16
Air Blank	0.000	23:17
Air Blank	0.000	23:19
Subject Sample # 2	0.211	23:20
Breath Volume	2.921	23:20
Air Blank	0.000	23:21
Air Blank	0.000	23:23
Subject Sample # 3	0.228	23:23
Breath Volume	2.671	23:23
Air Blank	0.000	23:24
Control Test	0.076	23:24
Air Blank	0.000	23:25
Diagnostics Check	OK	23:25

Data Download Date/Time: 11/29/2017 09:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILDE, MATTHEW, B
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: WILDE, MATTHEW
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 09/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 19:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:42

SUBJECT'S LAST NAME: ;6360102253201090092

DATE/TIME OF TEST: 10/01/2017 20:06:00

Diagnostics Check	OK	20:02	
Air Blank	0.000	20:03	
Control Test	0.080	20:03	
Air Blank	0.000	20:04	
Subject Sample	# 1	20:04	Subject Test Refused
Breath Volume	0.000	20:04	
Air Blank	0.000	20:05	
Control Test	0.080	20:05	
Air Blank	0.000	20:05	
Diagnostics Check	OK	20:06	

Data Download Date/Time: 10/26/2017 09:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILDE, MATTHEW, B
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: FEENEY, ERIC
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 09/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 20:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:24

SUBJECT'S LAST NAME: DORION

DATE/TIME OF TEST: 10/01/2017 21:48:38

Diagnostics Check	OK	21:45	
Air Blank	0.000	21:46	
Control Test	0.081	21:46	
Air Blank	0.000	21:46	
Subject Sample	# 1	21:47	Subject Test Refused
Breath Volume	0.000	21:47	
Air Blank	0.000	21:47	
Control Test	0.080	21:48	
Air Blank	0.000	21:48	
Diagnostics Check	OK	21:48	

Data Download Date/Time: 10/26/2017 09:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NABULSI,T,D
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: NABULSI,T
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 09/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 00:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:06

SUBJECT'S LAST NAME: COX

DATE/TIME OF TEST: 10/02/2017 00:47:56

Diagnostics Check	OK	00:41
Air Blank	0.000	00:42
Control Test	0.081	00:42
Air Blank	0.000	00:42
Subject Sample # 1	0.151	00:43
Breath Volume	2.511	00:43
Air Blank	0.000	00:44
Air Blank	0.000	00:45
Subject Sample # 2	0.167	00:46
Breath Volume	3.367	00:46
Air Blank	0.000	00:47
Control Test	0.081	00:47
Air Blank	0.000	00:47
Diagnostics Check	OK	00:47

Data Download Date/Time: 10/26/2017 09:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILDE, MATTHEW, B
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: WILLIAMS, DANIEL
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 09/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 02:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:54

SUBJECT'S LAST NAME: GIZOWSKI

DATE/TIME OF TEST: 10/02/2017 03:29:20

Diagnostics Check		OK	03:17	
Air Blank		0.000	03:17	
Control Test		0.081	03:18	
Air Blank		0.000	03:18	
Subject Sample	# 1	0.207	03:21	Volume Not Met
Breath Volume		0.664	03:21	
Air Blank		0.000	03:22	
Air Blank		0.000	03:24	
Subject Sample	# 2	0.210	03:27	Volume Not Met
Breath Volume		0.527	03:27	
Air Blank		0.000	03:28	
Control Test		0.079	03:28	
Air Blank		0.000	03:29	
Diagnostics Check		OK	03:29	

Data Download Date/Time: 10/26/2017 09:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILDE, MATTHEW, B
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: WILLIAMS, DANIEL
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 09/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 02:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:54

SUBJECT'S LAST NAME: GIZOWSKI

DATE/TIME OF TEST: 10/02/2017 03:43:20

Diagnostics Check		OK	03:31	
Air Blank		0.000	03:32	
Control Test		0.079	03:32	
Air Blank		0.000	03:32	
Subject Sample	# 1	0.210	03:36	Volume Not Met
Breath Volume		0.105	03:36	
Air Blank		0.000	03:36	
Air Blank		0.000	03:38	
Subject Sample	# 2	0.172	03:41	Volume Not Met
Breath Volume		0.097	03:41	
Air Blank		0.000	03:42	
Control Test		0.078	03:42	
Air Blank		0.000	03:43	
Diagnostics Check		OK	03:43	

Data Download Date/Time: 10/26/2017 09:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PITMAN,AARON,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: PITMAN,AARON
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 09/21/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 17:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:35

SUBJECT'S LAST NAME: KLAMERUS

DATE/TIME OF TEST: 10/07/2017 18:04:57

Diagnostics Check	OK	18:02	
Air Blank	0.000	18:02	
Control Test	0.081	18:02	
Air Blank	0.000	18:03	
Subject Sample	# 1	18:03	Subject Test Refused
Breath Volume	0.000	18:03	
Air Blank	0.000	18:04	
Control Test	0.081	18:04	
Air Blank	0.000	18:04	
Diagnostics Check	OK	18:04	

Data Download Date/Time: 10/26/2017 09:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FREMBGEN,TIA,A
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: CLAUSNER,ALEX
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 09/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 19:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:22

SUBJECT'S LAST NAME: HEBERT

DATE/TIME OF TEST: 10/07/2017 20:49:43

Diagnostics Check	OK	20:43
Air Blank	0.000	20:43
Control Test	0.081	20:44
Air Blank	0.000	20:44
Subject Sample # 1	0.155	20:45
Breath Volume	2.121	20:45
Air Blank	0.000	20:45
Air Blank	0.000	20:47
Subject Sample # 2	0.146	20:48
Breath Volume	1.851	20:48
Air Blank	0.000	20:48
Control Test	0.080	20:49
Air Blank	0.000	20:49
Diagnostics Check	OK	20:49

Data Download Date/Time: 10/26/2017 09:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NEAU,MICHAEL,D
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: NEAU,MICHAEL
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 09/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 22:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: JORDON

DATE/TIME OF TEST: 10/07/2017 23:17:34

Diagnostics Check	OK	23:14	
Air Blank	0.000	23:14	
Control Test	0.081	23:15	
Air Blank	0.000	23:15	
Subject Sample	# 1	23:16	Subject Test Refused
Breath Volume	0.000	23:16	
Air Blank	0.000	23:16	
Control Test	0.081	23:16	
Air Blank	0.000	23:17	
Diagnostics Check	OK	23:17	

Data Download Date/Time: 10/26/2017 09:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILHELM, TONY, J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: HOWARD, JODY
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 09/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 12:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: HILL

DATE/TIME OF TEST: 10/10/2017 01:37:06

Diagnostics Check	OK	01:31
Air Blank	0.000	01:31
Control Test	0.082	01:31
Air Blank	0.000	01:32
Subject Sample # 1	0.147	01:32
Breath Volume	3.105	01:32
Air Blank	0.000	01:33
Air Blank	0.000	01:35
Subject Sample # 2	0.147	01:35
Breath Volume	2.890	01:35
Air Blank	0.000	01:36
Control Test	0.080	01:36
Air Blank	0.000	01:37
Diagnostics Check	OK	01:37

Data Download Date/Time: 10/26/2017 09:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILDE, MATTHEW, B
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: WILDE, MATTHEW
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 09/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 02:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:06

SUBJECT'S LAST NAME: COLLIGNON

DATE/TIME OF TEST: 10/11/2017 03:32:05

Diagnostics Check	OK	03:28	
Air Blank	0.000	03:29	
Control Test	0.082	03:29	
Air Blank	0.000	03:30	
Subject Sample	# 1	03:30	Subject Test Refused
Breath Volume	0.000	03:30	
Air Blank	0.000	03:31	
Control Test	0.081	03:31	
Air Blank	0.000	03:31	
Diagnostics Check	OK	03:32	

Data Download Date/Time: 10/26/2017 09:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHELBY,MISTY,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: WILLIAMS,LANCE
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 09/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 19:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:14

SUBJECT'S LAST NAME: SHELBY

DATE/TIME OF TEST: 10/11/2017 20:40:38

Diagnostics Check	OK	20:35	
Air Blank	0.000	20:35	
Control Test	0.081	20:35	
Air Blank	0.000	20:36	
Subject Sample	# 1	20:39	Subject Test Refused
Breath Volume	0.000	20:39	
Air Blank	0.000	20:39	
Control Test	0.081	20:40	
Air Blank	0.000	20:40	
Diagnostics Check	OK	20:40	

Data Download Date/Time: 10/26/2017 09:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,LANCE,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: CLARK,MICHAEL
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 09/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 21:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:57

SUBJECT'S LAST NAME: DAURIA

DATE/TIME OF TEST: 10/11/2017 22:28:49

Diagnostics Check	OK	22:17	
Air Blank	0.000	22:17	
Control Test	0.081	22:18	
Air Blank	0.000	22:18	
Subject Sample	# 1	22:21	No Sample Provided
Breath Volume	0.000	22:21	
Air Blank	0.000	22:22	
Air Blank	0.000	22:24	
Subject Sample	# 2	22:27	No Sample Provided
Breath Volume	0.000	22:27	
Air Blank	0.000	22:27	
Control Test	0.080	22:28	
Air Blank	0.000	22:28	
Diagnostics Check	OK	22:28	

Data Download Date/Time: 10/26/2017 09:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUH,CASEY,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: DUH,CASEY
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 09/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 00:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: GEORGES

DATE/TIME OF TEST: 10/14/2017 01:55:47

Diagnostics Check	OK	01:52	
Air Blank	0.000	01:53	
Control Test	0.081	01:53	
Air Blank	0.000	01:54	
Subject Sample	# 1	01:54	Subject Test Refused
Breath Volume	0.000	01:54	
Air Blank	0.000	01:54	
Control Test	0.082	01:55	
Air Blank	0.000	01:55	
Diagnostics Check	OK	01:55	

Data Download Date/Time: 10/26/2017 09:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILHELM,TONY,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: HOOKS,JERRY
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 09/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 03:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: OVIEDO

DATE/TIME OF TEST: 10/14/2017 04:33:21

Diagnostics Check	OK	04:30	
Air Blank	0.000	04:30	
Control Test	0.081	04:31	
Air Blank	0.000	04:31	
Subject Sample	# 1	04:31	Subject Test Refused
Breath Volume	0.000	04:31	
Air Blank	0.000	04:32	
Control Test	0.081	04:32	
Air Blank	0.000	04:33	
Diagnostics Check	OK	04:33	

Data Download Date/Time: 10/26/2017 09:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICHMOND,BRENDAN,B
BREATH TEST OPERATOR AGENCY: Hurlburt Field PD
ARRESTING OFFICER: RICHMOND,BRENDAN
ARRESTING OFFICER AGENCY: Hurlburt Field PD
LAST AGENCY INSPECTION: 09/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 03:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: VAUGHAN

DATE/TIME OF TEST: 10/15/2017 04:48:34

Diagnostics Check	OK	04:42
Air Blank	0.000	04:42
Control Test	0.082	04:43
Air Blank	0.000	04:43
Subject Sample # 1	0.017	04:44
Breath Volume	2.554	04:44
Air Blank	0.000	04:44
Air Blank	0.000	04:46
Subject Sample # 2	0.016	04:47
Breath Volume	3.503	04:47
Air Blank	0.000	04:47
Control Test	0.080	04:47
Air Blank	0.000	04:48
Diagnostics Check	OK	04:48

Data Download Date/Time: 10/26/2017 09:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLARK,JEREMY,G
BREATH TEST OPERATOR AGENCY: Hurlburt Field PD
ARRESTING OFFICER: CLARK,JEREMY
ARRESTING OFFICER AGENCY: Hurlburt Field PD
LAST AGENCY INSPECTION: 09/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 04:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:27

SUBJECT'S LAST NAME: DANIELS

DATE/TIME OF TEST: 10/15/2017 07:19:30

Diagnostics Check	OK	07:12
Air Blank	0.000	07:12
Control Test	0.081	07:13
Air Blank	0.000	07:13
Subject Sample # 1	0.069	07:14
Breath Volume	2.437	07:14
Air Blank	0.000	07:15
Air Blank	0.000	07:17
Subject Sample # 2	0.070	07:17
Breath Volume	2.730	07:17
Air Blank	0.000	07:18
Control Test	0.079	07:18
Air Blank	0.000	07:19
Diagnostics Check	OK	07:19

Data Download Date/Time: 10/26/2017 09:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHAPMAN,AARON,C
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: CHAPMAN,AARON
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 09/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 15:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:00

SUBJECT'S LAST NAME: ARCHER

DATE/TIME OF TEST: 10/15/2017 16:37:31

Diagnostics Check	OK	16:21	
Air Blank	0.000	16:22	
Control Test	0.081	16:22	
Air Blank	0.000	16:23	
Subject Sample	# 1 0.208	16:24	
Breath Volume	1.480	16:24	
Air Blank	0.000	16:25	
Air Blank	0.000	16:27	
Subject Sample	# 2 0.175	16:30	Volume Not Met
Breath Volume	0.433	16:30	
Air Blank	0.000	16:31	
Air Blank	0.000	16:32	
Subject Sample	# 3 0.205	16:35	
Breath Volume	2.339	16:35	
Air Blank	0.000	16:36	
Control Test	0.075	16:36	
Air Blank	0.000	16:37	
Diagnostics Check	OK	16:37	

Data Download Date/Time: 10/26/2017 09:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILHELM,TONY,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: LECOCQ,MATTHEW
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 09/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 20:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:12

SUBJECT'S LAST NAME: ALEMAN-HERNANDEZ

DATE/TIME OF TEST: 10/15/2017 21:39:13

Diagnostics Check	OK	21:33
Air Blank	0.000	21:33
Control Test	0.081	21:33
Air Blank	0.000	21:34
Subject Sample # 1	0.251	21:34
Breath Volume	2.164	21:34
Air Blank	0.000	21:35
Air Blank	0.000	21:37
Subject Sample # 2	0.257	21:37
Breath Volume	2.617	21:37
Air Blank	0.000	21:38
Control Test	0.080	21:38
Air Blank	0.000	21:39
Diagnostics Check	OK	21:39

Data Download Date/Time: 10/26/2017 09:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POLAZZO, MICHAEL, A
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: POLAZZO, MICHAEL
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 09/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 20:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:53

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 10/15/2017 22:23:02

Diagnostics Check	OK	22:20	
Air Blank	0.000	22:20	
Control Test	0.080	22:20	
Air Blank	0.000	22:21	
Subject Sample	# 1	22:21	Subject Test Refused
Breath Volume	0.000	22:21	
Air Blank	0.000	22:22	
Control Test	0.080	22:22	
Air Blank	0.000	22:22	
Diagnostics Check	OK	22:23	

Data Download Date/Time: 10/26/2017 09:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRUJILLO, ANTHONY, R
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: VEGA, HECTOR
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 09/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 12:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 12:50

SUBJECT'S LAST NAME: COOLEY

DATE/TIME OF TEST: 10/19/2017 13:16:10

Diagnostics Check	OK	13:13	
Air Blank	0.000	13:13	
Control Test	0.082	13:13	
Air Blank	0.000	13:14	
Subject Sample	# 1	13:14	Subject Test Refused
Breath Volume	0.000	13:14	
Air Blank	0.000	13:15	
Control Test	0.081	13:15	
Air Blank	0.000	13:16	
Diagnostics Check	OK	13:16	

Data Download Date/Time: 10/26/2017 09:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,LANCE,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: WILLIAMS,LANCE
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 09/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 20:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:53

SUBJECT'S LAST NAME: ADAMS

DATE/TIME OF TEST: 10/22/2017 21:22:15

Diagnostics Check	OK	21:16
Air Blank	0.000	21:16
Control Test	0.082	21:17
Air Blank	0.000	21:17
Subject Sample # 1	0.000	21:17
Breath Volume	2.281	21:17
Air Blank	0.000	21:18
Air Blank	0.000	21:20
Subject Sample # 2	0.000	21:20
Breath Volume	2.093	21:20
Air Blank	0.000	21:21
Control Test	0.079	21:21
Air Blank	0.000	21:22
Diagnostics Check	OK	21:22

Data Download Date/Time: 10/26/2017 09:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILDE, MATTHEW, B
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: WILDE, MATTHEW
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 09/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 00:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:59

SUBJECT'S LAST NAME: BEMIS

DATE/TIME OF TEST: 10/25/2017 01:27:09

Diagnostics Check	OK	01:19
Air Blank	0.000	01:19
Control Test	0.082	01:20
Air Blank	0.000	01:20
Subject Sample # 1	0.071	01:22
Breath Volume	3.949	01:22
Air Blank	0.000	01:22
Air Blank	0.000	01:24
Subject Sample # 2	0.070	01:25
Breath Volume	2.929	01:25
Air Blank	0.000	01:26
Control Test	0.082	01:26
Air Blank	0.000	01:27
Diagnostics Check	OK	01:27

Data Download Date/Time: 10/26/2017 09:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000773

INSTRUMENT REGISTERED TO: LAFAYETTE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIS,CLEVELAND,D
BREATH TEST OPERATOR AGENCY: Lafayette County SO
ARRESTING OFFICER: ZIRKLE,ROBERT
ARRESTING OFFICER AGENCY: Lafayette County SO
LAST AGENCY INSPECTION: 09/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23016080A3
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 16:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:03

SUBJECT'S LAST NAME: WILLIAMSON

DATE/TIME OF TEST: 10/05/2017 17:35:47

Diagnostics Check	OK	17:24	
Air Blank	0.000	17:24	
Control Test	0.079	17:25	
Air Blank	0.000	17:25	
Subject Sample	# 1 0.304	17:28	Volume Not Met
Breath Volume	0.425	17:28	
Air Blank	0.000	17:29	
Air Blank	0.000	17:31	
Subject Sample	# 2	17:34	No Sample Provided
Breath Volume	0.000	17:34	
Air Blank	0.000	17:34	
Control Test	0.074	17:35	Control Outside Tolerance
Air Blank	0.000	17:35	

Data Download Date/Time: 11/28/2017 11:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000773

INSTRUMENT REGISTERED TO: LAFAYETTE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEAS,ALFRED,A
BREATH TEST OPERATOR AGENCY: Lafayette County SO
ARRESTING OFFICER: DEAS,ALFRED
ARRESTING OFFICER AGENCY: Lafayette County SO
LAST AGENCY INSPECTION: 09/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23016080A3
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: SEYERANCE

DATE/TIME OF TEST: 10/22/2017 20:27:17

Diagnostics Check	OK	20:21
Air Blank	0.000	20:21
Control Test	0.078	20:22
Air Blank	0.000	20:22
Subject Sample # 1	0.180	20:22
Breath Volume	1.777	20:22
Air Blank	0.000	20:23
Air Blank	0.000	20:25
Subject Sample # 2	0.170	20:25
Breath Volume	1.542	20:25
Air Blank	0.000	20:26
Control Test	0.076	20:26
Air Blank	0.000	20:27
Diagnostics Check	OK	20:27

Data Download Date/Time: 11/28/2017 11:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALKER,SHELIA,
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: WALKER,SHELIA
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 09/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 15:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:44

SUBJECT'S LAST NAME: BLACK

DATE/TIME OF TEST: 10/01/2017 04:11:14

Diagnostics Check	OK	04:04
Air Blank	0.000	04:05
Control Test	0.079	04:05
Air Blank	0.000	04:05
Subject Sample # 1	0.216	04:06
Breath Volume	2.351	04:06
Air Blank	0.000	04:07
Air Blank	0.000	04:08
Subject Sample # 2	0.227	04:09
Breath Volume	2.320	04:09
Air Blank	0.000	04:10
Control Test	0.079	04:10
Air Blank	0.000	04:11
Diagnostics Check	OK	04:11

Data Download Date/Time: 10/31/2017 17:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARVEY,DAVID,A
BREATH TEST OPERATOR AGENCY: Columbia County SO
ARRESTING OFFICER: NIERENBERG,MICHAEL
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 09/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:08

SUBJECT'S LAST NAME: RANDALL

DATE/TIME OF TEST: 10/15/2017 00:36:09

Diagnostics Check	OK	00:29
Air Blank	0.000	00:30
Control Test	0.079	00:30
Air Blank	0.000	00:31
Subject Sample # 1	0.172	00:31
Breath Volume	2.984	00:31
Air Blank	0.000	00:32
Air Blank	0.000	00:34
Subject Sample # 2	0.169	00:34
Breath Volume	2.023	00:34
Air Blank	0.000	00:35
Control Test	0.078	00:35
Air Blank	0.000	00:36
Diagnostics Check	OK	00:36

Data Download Date/Time: 10/31/2017 17:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAGNER,DANIEL,J
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: WAGNER,DANIEL
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 09/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: LUIZ

DATE/TIME OF TEST: 10/16/2017 23:51:26

Diagnostics Check	OK	23:45
Air Blank	0.000	23:45
Control Test	0.079	23:45
Air Blank	0.000	23:46
Subject Sample # 1	0.126	23:46
Breath Volume	3.664	23:46
Air Blank	0.000	23:47
Air Blank	0.000	23:49
Subject Sample # 2	0.128	23:49
Breath Volume	1.226	23:49
Air Blank	0.000	23:50
Control Test	0.079	23:50
Air Blank	0.000	23:51
Diagnostics Check	OK	23:51

Data Download Date/Time: 10/31/2017 17:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000780

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMSON,JACOB,L
BREATH TEST OPERATOR AGENCY: Suwannee County SO
ARRESTING OFFICER: WILLIAMSON,JACOB
ARRESTING OFFICER AGENCY: Suwannee County SO
LAST AGENCY INSPECTION: 09/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06616080A2
DRY GAS EXPIRY DATE: 04/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: DARCY

DATE/TIME OF TEST: 10/06/2017 02:41:36

Diagnostics Check	OK	02:33
Air Blank	0.000	02:33
Control Test	0.081	02:33
Air Blank	0.000	02:34
Subject Sample # 1	0.189	02:36
Breath Volume	3.648	02:36
Air Blank	0.000	02:37
Air Blank	0.000	02:38
Subject Sample # 2	0.191	02:39
Breath Volume	3.386	02:39
Air Blank	0.000	02:40
Control Test	0.080	02:40
Air Blank	0.000	02:41
Diagnostics Check	OK	02:41

Data Download Date/Time: 10/31/2017 21:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000780

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIEGEL,TIMOTHY,J
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: RIEGEL,TIMOTHY
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 09/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06616080A2
DRY GAS EXPIRY DATE: 04/05/2018
VIOLATION CODE: DUI
ARREST TIME: 15:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:40

SUBJECT'S LAST NAME: GAMBRELL

DATE/TIME OF TEST: 10/09/2017 17:04:57

Diagnostics Check	OK	17:01	
Air Blank	0.000	17:01	
Control Test	0.080	17:02	
Air Blank	0.000	17:02	
Subject Sample	# 1	17:03	RFI Detect
Breath Volume	0.542	17:03	
Air Blank	0.000	17:04	
Control Test	0.082	17:04	
Air Blank	0.000	17:04	
Diagnostics Check	OK	17:04	

Data Download Date/Time: 10/31/2017 21:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000780

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIEGEL,TIMOTHY,J
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: RIEGEL,TIMOTHY
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 09/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06616080A2
DRY GAS EXPIRY DATE: 04/05/2018
VIOLATION CODE: DUI
ARREST TIME: 15:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:40

SUBJECT'S LAST NAME: GAMBRELL

DATE/TIME OF TEST: 10/09/2017 17:14:21

Diagnostics Check	OK	17:07
Air Blank	0.000	17:08
Control Test	0.080	17:08
Air Blank	0.000	17:09
Subject Sample # 1	0.294	17:09
Breath Volume	3.800	17:09
Air Blank	0.000	17:10
Air Blank	0.000	17:12
Subject Sample # 2	0.291	17:12
Breath Volume	3.187	17:12
Air Blank	0.000	17:13
Control Test	0.080	17:13
Air Blank	0.000	17:14
Diagnostics Check	OK	17:14

Data Download Date/Time: 10/31/2017 21:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000780

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SLAUGHTER,DEREK,F
BREATH TEST OPERATOR AGENCY: Live Oak PD
ARRESTING OFFICER: SLAUGHTER,DEREK
ARRESTING OFFICER AGENCY: Live Oak PD
LAST AGENCY INSPECTION: 09/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06616080A2
DRY GAS EXPIRY DATE: 04/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:14

SUBJECT'S LAST NAME: RAMIREZ

DATE/TIME OF TEST: 10/14/2017 00:52:57

Diagnostics Check	OK	00:50	
Air Blank	0.000	00:50	
Control Test	0.081	00:50	
Air Blank	0.000	00:51	
Subject Sample	# 1	00:51	Subject Test Refused
Breath Volume	0.000	00:51	
Air Blank	0.000	00:52	
Control Test	0.081	00:52	
Air Blank	0.000	00:52	
Diagnostics Check	OK	00:52	

Data Download Date/Time: 10/31/2017 21:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000780

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOLIN,ALBERT,L
BREATH TEST OPERATOR AGENCY: FL Dept of AG
ARRESTING OFFICER: NOBLES,THAD
ARRESTING OFFICER AGENCY: FL Dept of AG
LAST AGENCY INSPECTION: 09/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06616080A2
DRY GAS EXPIRY DATE: 04/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: WESTER

DATE/TIME OF TEST: 10/14/2017 02:24:03

Diagnostics Check	OK	02:22	
Air Blank	0.000	02:23	
Control Test	0.074	02:23	Control Outside Tolerance
Air Blank	0.000	02:24	

Data Download Date/Time: 10/31/2017 21:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000780

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOLIN,ALBERT,L
BREATH TEST OPERATOR AGENCY: FL Dept of AG
ARRESTING OFFICER: NOBLES,THAD
ARRESTING OFFICER AGENCY: FL Dept of AG
LAST AGENCY INSPECTION: 09/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06616080A2
DRY GAS EXPIRY DATE: 04/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: WESTER

DATE/TIME OF TEST: 10/14/2017 02:32:02

Diagnostics Check	OK	02:26	
Air Blank	0.000	02:26	
Control Test	0.081	02:26	
Air Blank	0.000	02:27	
Subject Sample	# 1 0.209	02:30	Volume Not Met
Breath Volume	0.906	02:30	
Air Blank		02:31	Purge Fail
Air Blank		02:32	Purge Fail

Data Download Date/Time: 10/31/2017 21:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000780

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOLIN,ALBERT,L
BREATH TEST OPERATOR AGENCY: FL Dept of AG
ARRESTING OFFICER: NOBLES,THAD
ARRESTING OFFICER AGENCY: FL Dept of AG
LAST AGENCY INSPECTION: 09/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06616080A2
DRY GAS EXPIRY DATE: 04/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: WESTER

DATE/TIME OF TEST: 10/14/2017 02:42:08

Diagnostics Check	OK	02:38	
Air Blank	0.000	02:38	
Control Test	0.078	02:39	
Air Blank	0.000	02:39	
Subject Sample	# 1 0.226	02:40	
Breath Volume	1.648	02:40	
Air Blank		02:41	Purge Fail
Air Blank	0.000	02:42	

Data Download Date/Time: 10/31/2017 21:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000780

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOLIN,ALBERT,L
BREATH TEST OPERATOR AGENCY: FL Dept of AG
ARRESTING OFFICER: NOBLES,THAD
ARRESTING OFFICER AGENCY: FL Dept of AG
LAST AGENCY INSPECTION: 09/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06616080A2
DRY GAS EXPIRY DATE: 04/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: WESTER

DATE/TIME OF TEST: 10/14/2017 02:49:51

Diagnostics Check	OK	02:43	
Air Blank	0.000	02:44	
Control Test	0.078	02:44	
Air Blank	0.000	02:45	
Subject Sample	# 1 0.207	02:48	Volume Not Met
Breath Volume	0.304	02:48	
Air Blank		02:49	Purge Fail
Air Blank		02:49	Purge Fail

Data Download Date/Time: 10/31/2017 21:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000780

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLIE,JOHN,R
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: TILLIE,JOHN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 09/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06616080A2
DRY GAS EXPIRY DATE: 04/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: GREEN

DATE/TIME OF TEST: 10/18/2017 20:40:01

Diagnostics Check	OK	20:34
Air Blank	0.000	20:34
Control Test	0.082	20:34
Air Blank	0.000	20:35
Subject Sample # 1	0.000	20:35
Breath Volume	2.695	20:35
Air Blank	0.000	20:36
Air Blank	0.000	20:38
Subject Sample # 2	0.000	20:38
Breath Volume	2.445	20:38
Air Blank	0.000	20:39
Control Test	0.080	20:39
Air Blank	0.000	20:39
Diagnostics Check	OK	20:40

Data Download Date/Time: 10/31/2017 21:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000780

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,ALEXIS,
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: SINGLETARY,MATHEW
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 09/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06616080A2
DRY GAS EXPIRY DATE: 04/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 10/23/2017 22:57:06

Diagnostics Check	OK	22:50
Air Blank	0.000	22:51
Control Test	0.081	22:51
Air Blank	0.000	22:51
Subject Sample # 1	0.188	22:52
Breath Volume	2.566	22:52
Air Blank	0.000	22:53
Air Blank	0.000	22:54
Subject Sample # 2	0.183	22:55
Breath Volume	2.777	22:55
Air Blank	0.000	22:56
Control Test	0.079	22:56
Air Blank	0.000	22:57
Diagnostics Check	OK	22:57

Data Download Date/Time: 10/31/2017 21:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000780

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUGHES,ALFRED,L
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: HUGHES,ALFRED
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 09/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06616080A2
DRY GAS EXPIRY DATE: 04/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:31

SUBJECT'S LAST NAME: WINTHROP

DATE/TIME OF TEST: 10/25/2017 02:01:35

Diagnostics Check		OK	01:55
Air Blank		0.000	01:55
Control Test		0.081	01:56
Air Blank		0.000	01:56
Subject Sample	# 1	0.209	01:57
Breath Volume		1.210	01:57
Air Blank		0.000	01:57
Air Blank		0.000	01:59
Subject Sample	# 2	0.201	01:59
Breath Volume		1.589	01:59
Air Blank		0.000	02:00
Control Test		0.081	02:00
Air Blank		0.000	02:01
Diagnostics Check		OK	02:01

Data Download Date/Time: 10/31/2017 21:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000780

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SLAUGHTER,D,F
BREATH TEST OPERATOR AGENCY: Live Oak PD
ARRESTING OFFICER: SLAUGHTER,D
ARRESTING OFFICER AGENCY: Live Oak PD
LAST AGENCY INSPECTION: 09/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06616080A2
DRY GAS EXPIRY DATE: 04/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:36

SUBJECT'S LAST NAME: RIVAS

DATE/TIME OF TEST: 10/28/2017 03:11:13

Diagnostics Check		OK	03:04
Air Blank		0.000	03:05
Control Test		0.081	03:05
Air Blank		0.000	03:05
Subject Sample	# 1	0.231	03:06
Breath Volume		1.660	03:06
Air Blank		0.000	03:07
Air Blank		0.000	03:08
Subject Sample	# 2	0.245	03:09
Breath Volume		2.609	03:09
Air Blank		0.000	03:10
Control Test		0.079	03:10
Air Blank		0.000	03:11
Diagnostics Check		OK	03:11

Data Download Date/Time: 10/31/2017 21:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000781

INSTRUMENT REGISTERED TO: WASHINGTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LERNER, MATTHEW, H
BREATH TEST OPERATOR AGENCY: Chipley PD
ARRESTING OFFICER: FRENCH, CURTIS
ARRESTING OFFICER AGENCY: Washington County SO
LAST AGENCY INSPECTION: 09/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A1
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 17:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:22

SUBJECT'S LAST NAME: KEEGAN

DATE/TIME OF TEST: 10/02/2017 18:02:58

Diagnostics Check	OK	17:51	
Air Blank	0.000	17:51	
Control Test	0.077	17:52	
Air Blank	0.000	17:52	
Subject Sample	# 1 0.155	17:55	Volume Not Met
Breath Volume	0.117	17:55	
Air Blank	0.000	17:56	
Air Blank	0.000	17:58	
Subject Sample	# 2	18:01	No Sample Provided
Breath Volume	0.000	18:01	
Air Blank	0.000	18:02	
Control Test	0.076	18:02	
Air Blank	0.000	18:02	
Diagnostics Check	OK	18:02	

Data Download Date/Time: 10/03/2017 21:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000781

INSTRUMENT REGISTERED TO: WASHINGTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LERNER, MATTHEW, L
BREATH TEST OPERATOR AGENCY: Chipley PD
ARRESTING OFFICER: FRENCH, CURTIS
ARRESTING OFFICER AGENCY: Washington County SO
LAST AGENCY INSPECTION: 09/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A1
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 17:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:22

SUBJECT'S LAST NAME: KEEGAN

DATE/TIME OF TEST: 10/02/2017 18:17:49

Diagnostics Check	OK	18:06	
Air Blank	0.000	18:06	
Control Test	0.076	18:07	
Air Blank	0.000	18:07	
Subject Sample	# 1 0.166	18:10	Volume Not Met
Breath Volume	0.015	18:10	
Air Blank	0.000	18:11	
Air Blank	0.000	18:13	
Subject Sample	# 2	18:16	No Sample Provided
Breath Volume	0.000	18:16	
Air Blank	0.000	18:16	
Control Test	0.075	18:17	
Air Blank	0.000	18:17	
Diagnostics Check	OK	18:17	

Data Download Date/Time: 10/03/2017 21:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000781

INSTRUMENT REGISTERED TO: WASHINGTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRENCH,JASON,K
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: FRENCH,JASON
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 10/03/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A1
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 06:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:12

SUBJECT'S LAST NAME: CARTER

DATE/TIME OF TEST: 10/11/2017 08:40:18

Diagnostics Check	OK	08:33
Air Blank	0.000	08:34
Control Test	0.077	08:34
Air Blank	0.000	08:35
Subject Sample # 1	0.000	08:35
Breath Volume	1.437	08:35
Air Blank	0.000	08:36
Air Blank	0.000	08:38
Subject Sample # 2	0.000	08:38
Breath Volume	1.515	08:38
Air Blank	0.000	08:39
Control Test	0.076	08:39
Air Blank	0.000	08:40
Diagnostics Check	OK	08:40

Data Download Date/Time: 10/20/2017 12:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BANKS,SEAN,C
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: YOUNGREN,DENNIS
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 09/14/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: COX

DATE/TIME OF TEST: 10/05/2017 21:47:39

Diagnostics Check	OK	21:40
Air Blank	0.000	21:41
Control Test	0.080	21:41
Air Blank	0.000	21:42
Subject Sample # 1	0.047	21:43
Breath Volume	1.308	21:43
Air Blank	0.000	21:43
Air Blank	0.000	21:45
Subject Sample # 2	0.049	21:46
Breath Volume	2.796	21:46
Air Blank	0.000	21:46
Control Test	0.082	21:47
Air Blank	0.000	21:47
Diagnostics Check	OK	21:47

Data Download Date/Time: 10/17/2017 01:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCAPIN,BRANDON,L
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: SCAPIN,BRANDON
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 09/14/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:51

SUBJECT'S LAST NAME: GILROY

DATE/TIME OF TEST: 10/08/2017 01:31:34

Diagnostics Check	OK	01:28	
Air Blank	0.000	01:29	
Control Test	0.081	01:29	
Air Blank	0.000	01:30	
Subject Sample	# 1	01:30	Subject Test Refused
Breath Volume	0.000	01:30	
Air Blank	0.000	01:30	
Control Test	0.082	01:30	
Air Blank	0.000	01:31	
Diagnostics Check	OK	01:31	

Data Download Date/Time: 10/17/2017 01:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEIERS,BLAINE,T
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: WEIERS,BLAINE
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 09/14/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 18:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:31

SUBJECT'S LAST NAME: LAW

DATE/TIME OF TEST: 10/11/2017 19:07:21

Diagnostics Check		OK	18:52	
Air Blank		0.000	18:52	
Control Test		0.080	18:53	
Air Blank		0.000	18:53	
Subject Sample	# 1	0.000	18:56	Volume Not Met
Breath Volume		0.902	18:56	
Air Blank		0.000	18:57	
Air Blank		0.000	18:59	
Subject Sample	# 2	0.000	19:00	
Breath Volume		1.175	19:00	
Air Blank		0.000	19:00	
Air Blank		0.000	19:02	
Subject Sample	# 3	0.000	19:05	Volume Not Met
Breath Volume		0.539	19:05	
Air Blank		0.000	19:06	
Control Test		0.081	19:06	
Air Blank		0.000	19:07	
Diagnostics Check		OK	19:07	

Data Download Date/Time: 10/17/2017 01:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEIERS, BLAINE, T
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: WEIERS, BLAINE
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 09/14/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:11

SUBJECT'S LAST NAME: REED

DATE/TIME OF TEST: 10/13/2017 04:37:22

Diagnostics Check	OK	04:32	
Air Blank	0.000	04:32	
Control Test	0.080	04:32	
Air Blank	0.000	04:33	
Subject Sample # 1	0.209	04:36	
Breath Volume	1.980	04:36	
Air Blank		04:36	Purge Fail
Air Blank	0.000	04:37	

Data Download Date/Time: 10/17/2017 01:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEIERS,BLAINE,T
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: WEIERS,BLAINE
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 09/14/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:11

SUBJECT'S LAST NAME: REED

DATE/TIME OF TEST: 10/13/2017 04:46:53

Diagnostics Check		OK	04:40
Air Blank		0.000	04:40
Control Test		0.079	04:40
Air Blank		0.000	04:41
Subject Sample	# 1	0.204	04:42
Breath Volume		2.453	04:42
Air Blank		0.000	04:42
Air Blank		0.000	04:44
Subject Sample	# 2	0.205	04:45
Breath Volume		1.218	04:45
Air Blank		0.000	04:45
Control Test		0.080	04:46
Air Blank		0.000	04:46
Diagnostics Check		OK	04:46

Data Download Date/Time: 10/17/2017 01:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BANKS,SEAN,C
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: BANKS,SEAN
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 09/14/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:39

SUBJECT'S LAST NAME: RINES

DATE/TIME OF TEST: 10/13/2017 23:11:25

Diagnostics Check		OK	22:59	
Air Blank		0.000	23:00	
Control Test		0.079	23:00	
Air Blank		0.000	23:00	
Subject Sample	# 1	0.144	23:04	Volume Not Met
Breath Volume		0.097	23:04	
Air Blank		0.000	23:04	
Air Blank		0.000	23:06	
Subject Sample	# 2	0.146	23:09	Volume Not Met
Breath Volume		1.222	23:09	
Air Blank		0.000	23:10	
Control Test		0.082	23:10	
Air Blank		0.000	23:11	
Diagnostics Check		OK	23:11	

Data Download Date/Time: 10/17/2017 01:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RINES,VALICIA,L
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: BANKS,SEAN
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 09/14/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:39

SUBJECT'S LAST NAME: RINES

DATE/TIME OF TEST: 10/13/2017 23:22:44

Diagnostics Check	OK	23:17	
Air Blank	0.000	23:18	
Control Test	0.080	23:18	
Air Blank	0.000	23:18	
Subject Sample # 1	0.177	23:21	
Breath Volume	1.890	23:21	
Air Blank		23:22	Purge Fail
Air Blank	0.000	23:22	

Data Download Date/Time: 10/17/2017 01:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERSON,THOMAS,C
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: XXXXX,XXXXXX
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 09/14/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: SYS CHECK
ARREST TIME: 13:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:00

SUBJECT'S LAST NAME: XXXXXXXX

DATE/TIME OF TEST: 10/14/2017 14:06:29

Diagnostics Check	OK	14:00
Air Blank	0.000	14:00
Control Test	0.080	14:01
Air Blank	0.000	14:01
Subject Sample # 1	0.000	14:02
Breath Volume	3.101	14:02
Air Blank	0.000	14:02
Air Blank	0.000	14:04
Subject Sample # 2	0.000	14:05
Breath Volume	2.652	14:05
Air Blank	0.000	14:05
Control Test	0.081	14:05
Air Blank	0.000	14:06
Diagnostics Check	OK	14:06

Data Download Date/Time: 10/17/2017 01:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERSON,THOMAS,C
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: TURNER,RICHARD
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 09/14/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:43

SUBJECT'S LAST NAME: HEITZ

DATE/TIME OF TEST: 10/16/2017 21:10:14

Diagnostics Check	OK	21:07	
Air Blank	0.000	21:07	
Control Test	0.081	21:07	
Air Blank	0.000	21:08	
Subject Sample	# 1	21:08	Subject Test Refused
Breath Volume	0.000	21:08	
Air Blank	0.000	21:09	
Control Test	0.081	21:09	
Air Blank	0.000	21:10	
Diagnostics Check	OK	21:10	

Data Download Date/Time: 10/17/2017 01:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNGREN,DENNIS,A
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: YOUNGREN,DENNIS
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 10/16/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:24

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 10/19/2017 00:48:07

Diagnostics Check	OK	00:47	
Air Blank		00:47	Ambient Fail
Air Blank	0.000	00:48	

Data Download Date/Time: 10/23/2017 06:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNGREN,DENNIS,A
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: YOUNGREN,DENNIS
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 10/16/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:24

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 10/19/2017 00:57:21

Diagnostics Check	OK	00:50
Air Blank	0.000	00:51
Control Test	0.081	00:51
Air Blank	0.000	00:51
Subject Sample # 1	0.101	00:52
Breath Volume	3.257	00:52
Air Blank	0.000	00:53
Air Blank	0.000	00:55
Subject Sample # 2	0.099	00:55
Breath Volume	3.332	00:55
Air Blank	0.000	00:56
Control Test	0.081	00:56
Air Blank	0.000	00:57
Diagnostics Check	OK	00:57

Data Download Date/Time: 10/23/2017 06:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERSON,THOMAS,C
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: WEIERS,BLAINE
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 10/16/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:37

SUBJECT'S LAST NAME: WOODS

DATE/TIME OF TEST: 10/21/2017 00:04:43

Diagnostics Check	OK	23:58
Air Blank	0.000	23:59
Control Test	0.081	23:59
Air Blank	0.000	23:59
Subject Sample # 1	0.050	00:00
Breath Volume	2.804	00:00
Air Blank	0.000	00:00
Air Blank	0.000	00:02
Subject Sample # 2	0.051	00:03
Breath Volume	3.742	00:03
Air Blank	0.000	00:03
Control Test	0.082	00:04
Air Blank	0.000	00:04
Diagnostics Check	OK	00:04

Data Download Date/Time: 10/23/2017 06:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PULLEY,PATRICK,S
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: SCAPIN,BRANDON
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 10/16/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 18:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:41

SUBJECT'S LAST NAME: GIBBS

DATE/TIME OF TEST: 10/22/2017 19:05:46

Diagnostics Check	OK	19:02	
Air Blank	0.000	19:02	
Control Test	0.081	19:02	
Air Blank	0.000	19:03	
Subject Sample	# 1	19:04	Subject Test Refused
Breath Volume	0.000	19:04	
Air Blank	0.000	19:04	
Control Test	0.080	19:05	
Air Blank	0.000	19:05	
Diagnostics Check	OK	19:05	

Data Download Date/Time: 10/23/2017 06:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNGREN,DENNIS,A
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: YOUNGREN,DENNIS
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 10/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:06

SUBJECT'S LAST NAME: MASTAW

DATE/TIME OF TEST: 10/28/2017 02:32:33

Diagnostics Check		OK	02:26
Air Blank		0.000	02:26
Control Test		0.080	02:27
Air Blank		0.000	02:27
Subject Sample	# 1	0.201	02:28
Breath Volume		2.019	02:28
Air Blank		0.000	02:28
Air Blank		0.000	02:30
Subject Sample	# 2	0.210	02:31
Breath Volume		1.375	02:31
Air Blank		0.000	02:31
Control Test		0.080	02:31
Air Blank		0.000	02:32
Diagnostics Check		OK	02:32

Data Download Date/Time: 10/30/2017 14:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNGREN,DENNIS,A
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: YOUNGREN,DENNIS
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 10/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 18:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:05

SUBJECT'S LAST NAME: PATEL

DATE/TIME OF TEST: 10/29/2017 19:49:33

Diagnostics Check	OK	19:43	
Air Blank	0.000	19:43	
Control Test	0.081	19:44	
Air Blank	0.000	19:44	
Subject Sample # 1	0.000	19:45	
Breath Volume	1.128	19:45	
Air Blank	0.000	19:45	
Air Blank	0.000	19:47	
Subject Sample # 2		19:48	Subject Test Refused
Breath Volume	0.000	19:48	
Air Blank	0.000	19:48	
Control Test	0.081	19:48	
Air Blank	0.000	19:49	
Diagnostics Check	OK	19:49	

Data Download Date/Time: 10/30/2017 14:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BANKS,SEAN,C
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: BANKS,SEAN
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 10/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:23

SUBJECT'S LAST NAME: DISINGER

DATE/TIME OF TEST: 10/30/2017 01:49:29

Diagnostics Check	OK	01:43
Air Blank	0.000	01:43
Control Test	0.080	01:44
Air Blank	0.000	01:44
Subject Sample # 1	0.172	01:45
Breath Volume	4.273	01:45
Air Blank	0.000	01:45
Air Blank	0.000	01:47
Subject Sample # 2	0.174	01:47
Breath Volume	4.367	01:47
Air Blank	0.000	01:48
Control Test	0.083	01:48
Air Blank	0.000	01:49
Diagnostics Check	OK	01:49

Data Download Date/Time: 10/30/2017 14:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000788

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DRAWDY,LAINE,A
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: DRAWDY,LAINE
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 10/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: BOST

DATE/TIME OF TEST: 10/13/2017 04:44:06

Diagnostics Check	OK	04:34
Air Blank	0.000	04:34
Control Test	0.081	04:35
Air Blank	0.000	04:35
Subject Sample # 1	0.152	04:36
Breath Volume	1.359	04:36
Air Blank	0.000	04:36
Air Blank	0.000	04:38
Subject Sample # 2	0.180	04:39
Breath Volume	3.179	04:39
Air Blank	0.000	04:40
Air Blank	0.000	04:41
Subject Sample # 3	0.185	04:42
Breath Volume	2.957	04:42
Air Blank	0.000	04:43
Control Test	0.080	04:43
Air Blank	0.000	04:44
Diagnostics Check	OK	04:44

Data Download Date/Time: 11/15/2017 11:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000788

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DRAWDY,LAINE,A
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: DRAWDY,LAINE
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 10/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 71817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:40

SUBJECT'S LAST NAME: STEWART

DATE/TIME OF TEST: 10/13/2017 22:05:56

Diagnostics Check	OK	21:59
Air Blank	0.000	22:00
Control Test	0.081	22:00
Air Blank	0.000	22:00
Subject Sample # 1	0.191	22:01
Breath Volume	1.617	22:01
Air Blank	0.000	22:02
Air Blank	0.000	22:03
Subject Sample # 2	0.191	22:04
Breath Volume	1.601	22:04
Air Blank	0.000	22:05
Control Test	0.081	22:05
Air Blank	0.000	22:05
Diagnostics Check	OK	22:05

Data Download Date/Time: 11/15/2017 11:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000788

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASSERLY,STEPHEN,R
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: ARCHBAULT,ERIK
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 10/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 0017
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: CLAYTON

DATE/TIME OF TEST: 10/14/2017 00:11:31

Diagnostics Check	OK	00:04
Air Blank	0.000	00:05
Control Test	0.081	00:05
Air Blank	0.000	00:06
Subject Sample # 1	0.224	00:06
Breath Volume	3.121	00:06
Air Blank	0.000	00:07
Air Blank	0.000	00:09
Subject Sample # 2	0.224	00:09
Breath Volume	2.683	00:09
Air Blank	0.000	00:10
Control Test	0.079	00:10
Air Blank	0.000	00:11
Diagnostics Check	OK	00:11

Data Download Date/Time: 11/15/2017 11:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000788

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,MICHAEL,L
BREATH TEST OPERATOR AGENCY: Fort Pierce PD
ARRESTING OFFICER: FRAZIER,MICHAEL
ARRESTING OFFICER AGENCY: Fort Pierce PD
LAST AGENCY INSPECTION: 10/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 80000788
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: BUTLER

DATE/TIME OF TEST: 10/15/2017 00:37:48

Diagnostics Check	OK	00:31
Air Blank	0.000	00:31
Control Test	0.082	00:32
Air Blank	0.000	00:32
Subject Sample # 1	0.184	00:33
Breath Volume	2.289	00:33
Air Blank	0.000	00:33
Air Blank	0.000	00:35
Subject Sample # 2	0.186	00:36
Breath Volume	1.656	00:36
Air Blank	0.000	00:36
Control Test	0.082	00:37
Air Blank	0.000	00:37
Diagnostics Check	OK	00:37

Data Download Date/Time: 11/15/2017 11:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000788

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DRAWDY,LAINE,A
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: DRAWDY,LAINE
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 10/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: GRANTHAM

DATE/TIME OF TEST: 10/15/2017 02:56:45

Diagnostics Check	OK	02:50
Air Blank	0.000	02:50
Control Test	0.081	02:51
Air Blank	0.000	02:51
Subject Sample # 1	0.120	02:52
Breath Volume	1.347	02:52
Air Blank	0.000	02:52
Air Blank	0.000	02:54
Subject Sample # 2	0.117	02:55
Breath Volume	1.203	02:55
Air Blank	0.000	02:55
Control Test	0.081	02:56
Air Blank	0.000	02:56
Diagnostics Check	OK	02:56

Data Download Date/Time: 11/15/2017 11:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000788

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,RICHARD,D
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: COX,TRENT
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 10/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817980A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 05:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:20

SUBJECT'S LAST NAME: RAY

DATE/TIME OF TEST: 10/16/2017 06:54:41

Diagnostics Check	OK	06:47
Air Blank	0.000	06:47
Control Test	0.081	06:48
Air Blank	0.000	06:48
Subject Sample # 1	0.097	06:49
Breath Volume	2.175	06:49
Air Blank	0.000	06:50
Air Blank	0.000	06:52
Subject Sample # 2	0.088	06:52
Breath Volume	1.582	06:52
Air Blank	0.000	06:53
Control Test	0.079	06:54
Air Blank	0.000	06:54
Diagnostics Check	OK	06:54

Data Download Date/Time: 11/15/2017 11:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000790

INSTRUMENT REGISTERED TO: VERO BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIS, JEREMIAH, T
BREATH TEST OPERATOR AGENCY: Vero Beach PD
ARRESTING OFFICER: WILLIS, JEREMIAH
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 09/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A6
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: HURST

DATE/TIME OF TEST: 10/08/2017 01:35:16

Diagnostics Check	OK	01:34	
Air Blank	0.000	01:34	
		01:34	RFI Detect
Air Blank	0.000	01:35	

Data Download Date/Time: 12/07/2017 16:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000790

INSTRUMENT REGISTERED TO: VERO BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIS, JEREMIAH, W
BREATH TEST OPERATOR AGENCY: Vero Beach PD
ARRESTING OFFICER: WILLIS, JEREMIAH
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 09/24/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 12716080A6
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: HURST

DATE/TIME OF TEST: 10/08/2017 01:50:03

Diagnostics Check		OK	01:38	
Air Blank		0.000	01:38	
Control Test		0.078	01:38	
Air Blank		0.000	01:39	
Subject Sample	# 1		01:42	No Sample Provided
Breath Volume		0.000	01:42	
Air Blank		0.000	01:43	
Air Blank		0.000	01:45	
Subject Sample	# 2	0.219	01:45	
Breath Volume		2.035	01:45	
Air Blank		0.000	01:46	
Air Blank		0.000	01:48	
Subject Sample	# 3	0.219	01:48	
Breath Volume		1.792	01:48	
Air Blank		0.000	01:49	
Control Test		0.077	01:49	
Air Blank		0.000	01:49	
Diagnostics Check		OK	01:50	

Data Download Date/Time: 12/07/2017 16:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000790

INSTRUMENT REGISTERED TO: VERO BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CROWLEY,SEAN,W
BREATH TEST OPERATOR AGENCY: Vero Beach PD
ARRESTING OFFICER: SOMOZA,EDUARDO
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 09/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A6
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: CARMICHAEL

DATE/TIME OF TEST: 10/23/2017 03:43:27

Diagnostics Check		OK	03:36
Air Blank		0.000	03:36
Control Test		0.079	03:37
Air Blank		0.000	03:37
Subject Sample	# 1	0.171	03:38
Breath Volume		2.152	03:38
Air Blank		0.000	03:38
Air Blank		0.000	03:40
Subject Sample	# 2	0.178	03:41
Breath Volume		2.101	03:41
Air Blank		0.000	03:42
Control Test		0.079	03:42
Air Blank		0.000	03:43
Diagnostics Check		OK	03:43

Data Download Date/Time: 12/07/2017 16:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000790

INSTRUMENT REGISTERED TO: VERO BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PLUMB,GREGORY,T
BREATH TEST OPERATOR AGENCY: Vero Beach PD
ARRESTING OFFICER: PLUMB,GREGORY
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 09/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A6
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:55

SUBJECT'S LAST NAME: BRYANT

DATE/TIME OF TEST: 10/24/2017 22:18:13

Diagnostics Check	OK	22:15	
Air Blank	0.000	22:15	
Control Test	0.079	22:16	
Air Blank	0.000	22:16	
Subject Sample	# 1	22:16	Subject Test Refused
Breath Volume	0.000	22:16	
Air Blank	0.000	22:17	
Control Test	0.079	22:17	
Air Blank	0.000	22:18	
Diagnostics Check	OK	22:18	

Data Download Date/Time: 12/07/2017 16:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARNER, TODD, A
BREATH TEST OPERATOR AGENCY: Fort Pierce PD
ARRESTING OFFICER: ECKRODE, ALBERT
ARRESTING OFFICER AGENCY: Fort Pierce PD
LAST AGENCY INSPECTION: 09/14/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 12:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:04

SUBJECT'S LAST NAME: OVER

DATE/TIME OF TEST: 10/03/2017 13:28:59

Diagnostics Check	OK	13:24	
Air Blank	0.000	13:25	
Control Test	0.080	13:25	
Air Blank	0.000	13:26	
Subject Sample	# 1	13:27	Subject Test Refused
Breath Volume	0.000	13:27	
Air Blank	0.000	13:28	
Control Test	0.080	13:28	
Air Blank	0.000	13:28	
Diagnostics Check	OK	13:28	

Data Download Date/Time: 10/11/2017 10:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DRAWDY,LAINE,A
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: DRAWDY,LAINE
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 09/14/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: STANTON

DATE/TIME OF TEST: 10/05/2017 01:09:15

Diagnostics Check	OK	01:00
Air Blank	0.000	01:00
Control Test	0.080	01:01
Air Blank	0.000	01:01
Subject Sample # 1	0.138	01:03
Breath Volume	1.390	01:03
Air Blank	0.000	01:04
Air Blank	0.000	01:06
Subject Sample # 2	0.137	01:07
Breath Volume	2.148	01:07
Air Blank	0.000	01:08
Control Test	0.079	01:08
Air Blank	0.000	01:09
Diagnostics Check	OK	01:09

Data Download Date/Time: 10/11/2017 10:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GULLETT,JUSTIN,T
BREATH TEST OPERATOR AGENCY: Fort Pierce PD
ARRESTING OFFICER: GULLETT,JUSTIN
ARRESTING OFFICER AGENCY: Fort Pierce PD
LAST AGENCY INSPECTION: 09/14/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 13:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:48

SUBJECT'S LAST NAME: STOCKTON

DATE/TIME OF TEST: 10/08/2017 14:37:35

Diagnostics Check	OK	14:31	
Air Blank	0.000	14:31	
Control Test	0.079	14:32	
Air Blank	0.000	14:32	
Subject Sample # 1	0.345	14:33	
Breath Volume	3.085	14:33	
Air Blank	0.000	14:33	
Air Blank	0.000	14:35	
Subject Sample # 2		14:35	Slope Not Met
Breath Volume	1.914	14:35	
Air Blank	0.000	14:36	
Control Test	0.078	14:36	
Air Blank	0.000	14:37	
Diagnostics Check	OK	14:37	

Data Download Date/Time: 10/11/2017 10:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GULLETT,JUSTIN,T
BREATH TEST OPERATOR AGENCY: Fort Pierce PD
ARRESTING OFFICER: GULLETT,JUSTIN
ARRESTING OFFICER AGENCY: Fort Pierce PD
LAST AGENCY INSPECTION: 09/14/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 13:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:48

SUBJECT'S LAST NAME: STOCKTON

DATE/TIME OF TEST: 10/08/2017 14:43:03

Diagnostics Check	OK	14:39	
Air Blank	0.000	14:40	
Control Test	0.079	14:40	
Air Blank	0.000	14:41	
Subject Sample # 1	0.333	14:41	
Breath Volume	2.488	14:41	
Air Blank		14:42	Purge Fail
Air Blank	0.000	14:43	

Data Download Date/Time: 10/11/2017 10:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WITT, TYLER,
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: WITT, TYLER
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 09/14/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:52

SUBJECT'S LAST NAME: MENENDEZ SARRASQUETA

DATE/TIME OF TEST: 10/11/2017 04:33:13

Diagnostics Check	OK	04:27
Air Blank	0.000	04:27
Control Test	0.080	04:28
Air Blank	0.000	04:28
Subject Sample # 1	0.124	04:28
Breath Volume	1.375	04:28
Air Blank	0.000	04:29
Air Blank	0.000	04:31
Subject Sample # 2	0.121	04:31
Breath Volume	1.597	04:31
Air Blank	0.000	04:32
Control Test	0.079	04:32
Air Blank	0.000	04:33
Diagnostics Check	OK	04:33

Data Download Date/Time: 10/11/2017 10:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERRING,TIMOTHY,R
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: HERRING,TIMTOHY
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 10/20/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:12

SUBJECT'S LAST NAME: CORNELL

DATE/TIME OF TEST: 10/20/2017 23:00:45

Diagnostics Check	OK	22:51
Air Blank	0.000	22:51
Control Test	0.079	22:52
Air Blank	0.000	22:52
Subject Sample # 1	0.208	22:55
Breath Volume	1.347	22:55
Air Blank	0.000	22:57
Air Blank	0.000	22:58
Subject Sample # 2	0.210	22:58
Breath Volume	1.476	22:58
Air Blank	0.000	22:59
Control Test	0.075	23:00
Air Blank	0.000	23:00
Diagnostics Check	OK	23:00

Data Download Date/Time: 11/09/2017 11:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DRAWDY,LAINE,A
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: JOYNER,TIMOTHY
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 10/20/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:22

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 10/21/2017 03:00:12

Diagnostics Check		OK	02:45	
Air Blank		0.000	02:46	
Control Test		0.080	02:46	
Air Blank		0.000	02:46	
Subject Sample	# 1	0.170	02:47	
Breath Volume		1.167	02:47	
Air Blank		0.000	02:47	
Air Blank		0.000	02:49	
Subject Sample	# 2	0.152	02:52	Volume Not Met
Breath Volume		0.160	02:52	
Air Blank		0.000	02:53	
Air Blank		0.000	02:55	
Subject Sample	# 3	0.155	02:58	Volume Not Met
Breath Volume		0.218	02:58	
Air Blank		0.000	02:59	
Control Test		0.079	02:59	
Air Blank		0.000	03:00	
Diagnostics Check		OK	03:00	

Data Download Date/Time: 11/09/2017 11:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAYNE,JAMES,B
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: SHERMETARO,MARTIN
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 10/20/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:27

SUBJECT'S LAST NAME: VALENTIN SILVA

DATE/TIME OF TEST: 10/21/2017 03:13:21

Diagnostics Check	OK	03:04
Air Blank	0.000	03:04
Control Test	0.079	03:05
Air Blank	0.000	03:05
Subject Sample # 1	0.145	03:08
Breath Volume	2.058	03:08
Air Blank	0.000	03:08
Air Blank	0.000	03:10
Subject Sample # 2	0.152	03:11
Breath Volume	2.734	03:11
Air Blank	0.000	03:12
Control Test	0.079	03:12
Air Blank	0.000	03:13
Diagnostics Check	OK	03:13

Data Download Date/Time: 11/09/2017 11:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DRAWDY,LAINE,A
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: DRAWDY,LAINE
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 10/20/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:19

SUBJECT'S LAST NAME: MEDOZA ALAVA

DATE/TIME OF TEST: 10/28/2017 00:48:36

Diagnostics Check		OK	00:40
Air Blank		0.000	00:40
Control Test		0.080	00:40
Air Blank		0.000	00:41
Subject Sample	# 1	0.104	00:42
Breath Volume		1.425	00:42
Air Blank		0.000	00:43
Air Blank		0.000	00:44
Subject Sample	# 2	0.110	00:47
Breath Volume		1.625	00:47
Air Blank		0.000	00:47
Control Test		0.079	00:47
Air Blank		0.000	00:48
Diagnostics Check		OK	00:48

Data Download Date/Time: 11/09/2017 11:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DRAWDY,LAINE,A
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: MONDRAGON,LIONEL
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 10/20/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: PEREIRA

DATE/TIME OF TEST: 10/28/2017 02:25:29

Diagnostics Check	OK	02:18
Air Blank	0.000	02:18
Control Test	0.079	02:18
Air Blank	0.000	02:19
Subject Sample # 1	0.234	02:20
Breath Volume	2.414	02:20
Air Blank	0.000	02:21
Air Blank	0.000	02:23
Subject Sample # 2	0.238	02:23
Breath Volume	1.824	02:23
Air Blank	0.000	02:24
Control Test	0.079	02:24
Air Blank	0.000	02:25
Diagnostics Check	OK	02:25

Data Download Date/Time: 11/09/2017 11:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STOKES,SHAYNE,R
BREATH TEST OPERATOR AGENCY: Fort Pierce PD
ARRESTING OFFICER: STOKES,SHAYNE
ARRESTING OFFICER AGENCY: Fort Pierce PD
LAST AGENCY INSPECTION: 10/20/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 16:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:19

SUBJECT'S LAST NAME: HALL

DATE/TIME OF TEST: 10/28/2017 16:51:18

Diagnostics Check		OK	16:39
Air Blank		0.000	16:40
Control Test		0.079	16:40
Air Blank		0.000	16:41
Subject Sample	# 1	0.255	16:42
Breath Volume		2.250	16:42
Air Blank		0.000	16:43
Air Blank		0.000	16:45
Subject Sample	# 2	0.286	16:46
Breath Volume		2.332	16:46
Air Blank		0.000	16:47
Air Blank		0.000	16:49
Subject Sample	# 3	0.245	16:49
Breath Volume		1.523	16:49
Air Blank		0.000	16:50
Control Test		0.077	16:50
Air Blank		0.000	16:51
Diagnostics Check		OK	16:51

Data Download Date/Time: 11/09/2017 11:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DRAWDY,LAINE,A
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: DRAWDY,LAINE
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 10/20/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 10/28/2017 23:57:07

Diagnostics Check		OK	23:50
Air Blank		0.000	23:50
Control Test		0.079	23:51
Air Blank		0.000	23:51
Subject Sample	# 1	0.285	23:52
Breath Volume		2.421	23:52
Air Blank		0.000	23:53
Air Blank		0.000	23:54
Subject Sample	# 2	0.277	23:55
Breath Volume		2.226	23:55
Air Blank		0.000	23:56
Control Test		0.078	23:56
Air Blank		0.000	23:57
Diagnostics Check		OK	23:57

Data Download Date/Time: 11/09/2017 11:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000797

INSTRUMENT REGISTERED TO: HAMILTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH JR,LORNE,J
BREATH TEST OPERATOR AGENCY: Hamilton County SO
ARRESTING OFFICER: SMITH JR,LORNE
ARRESTING OFFICER AGENCY: Hamilton County SO
LAST AGENCY INSPECTION: 09/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 18:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:05

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 10/05/2017 19:29:01

Diagnostics Check	OK	19:26	
Air Blank	0.000	19:26	
Control Test	0.079	19:26	
Air Blank	0.000	19:27	
Subject Sample	# 1	19:27	Subject Test Refused
Breath Volume	0.000	19:27	
Air Blank	0.000	19:28	
Control Test	0.079	19:28	
Air Blank	0.000	19:28	
Diagnostics Check	OK	19:29	

Data Download Date/Time: 10/31/2017 01:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000797

INSTRUMENT REGISTERED TO: HAMILTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH JR,LORNE,J
BREATH TEST OPERATOR AGENCY: Hamilton County SO
ARRESTING OFFICER: THOMAS,LARRY
ARRESTING OFFICER AGENCY: Hamilton County SO
LAST AGENCY INSPECTION: 09/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: ROSE

DATE/TIME OF TEST: 10/06/2017 23:50:34

Diagnostics Check	OK	23:47	
Air Blank	0.000	23:48	
Control Test	0.080	23:48	
Air Blank	0.000	23:48	
Subject Sample	# 1	23:49	Subject Test Refused
Breath Volume	0.000	23:49	
Air Blank	0.000	23:49	
Control Test	0.079	23:49	
Air Blank	0.000	23:50	
Diagnostics Check	OK	23:50	

Data Download Date/Time: 10/31/2017 01:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000797

INSTRUMENT REGISTERED TO: HAMILTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CURRY, WILLIAM, C
BREATH TEST OPERATOR AGENCY: Hamilton County SO
ARRESTING OFFICER: GARMANY, TY
ARRESTING OFFICER AGENCY: Jennings PD
LAST AGENCY INSPECTION: 09/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: DANIELS

DATE/TIME OF TEST: 10/12/2017 02:12:11

Diagnostics Check	OK	02:06
Air Blank	0.000	02:06
Control Test	0.080	02:06
Air Blank	0.000	02:07
Subject Sample # 1	0.042	02:07
Breath Volume	2.222	02:07
Air Blank	0.000	02:08
Air Blank	0.000	02:10
Subject Sample # 2	0.043	02:10
Breath Volume	1.500	02:10
Air Blank	0.000	02:11
Control Test	0.080	02:11
Air Blank	0.000	02:12
Diagnostics Check	OK	02:12

Data Download Date/Time: 10/31/2017 01:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000797

INSTRUMENT REGISTERED TO: HAMILTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDONALD,SAMUEL,W
BREATH TEST OPERATOR AGENCY: Jasper PD
ARRESTING OFFICER: NOBLES,THAD
ARRESTING OFFICER AGENCY: FL Dept of AG
LAST AGENCY INSPECTION: 09/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:43

SUBJECT'S LAST NAME: WESTER

DATE/TIME OF TEST: 10/14/2017 04:11:58

Diagnostics Check	OK	04:03	
Air Blank	0.000	04:03	
Control Test	0.080	04:04	
Air Blank	0.000	04:04	
Subject Sample	# 1	04:07	No Sample Provided
Breath Volume	0.000	04:07	
Air Blank	0.000	04:08	
Air Blank	0.000	04:10	
Subject Sample	# 2	04:10	Subject Test Refused
Breath Volume	0.000	04:10	
Air Blank	0.000	04:11	
Control Test	0.079	04:11	
Air Blank	0.000	04:11	
Diagnostics Check	OK	04:11	

Data Download Date/Time: 10/31/2017 01:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRIMBLE, TERESA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DEJESUS, V
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:53

SUBJECT'S LAST NAME: ELLANA

DATE/TIME OF TEST: 10/01/2017 22:19:57

Diagnostics Check	OK	22:13
Air Blank	0.000	22:13
Control Test	0.081	22:14
Air Blank	0.000	22:14
Subject Sample # 1	0.170	22:15
Breath Volume	1.636	22:15
Air Blank	0.000	22:16
Air Blank	0.000	22:17
Subject Sample # 2	0.171	22:18
Breath Volume	1.195	22:18
Air Blank	0.000	22:19
Control Test	0.080	22:19
Air Blank	0.000	22:19
Diagnostics Check	OK	22:19

Data Download Date/Time: 10/04/2017 19:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRENNAN, D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:56

SUBJECT'S LAST NAME: RINGER

DATE/TIME OF TEST: 10/01/2017 23:24:10

Diagnostics Check	OK	23:17
Air Blank	0.000	23:18
Control Test	0.080	23:18
Air Blank	0.000	23:18
Subject Sample # 1	0.317	23:19
Breath Volume	3.507	23:19
Air Blank	0.000	23:20
Air Blank	0.000	23:21
Subject Sample # 2	0.308	23:22
Breath Volume	2.929	23:22
Air Blank	0.000	23:23
Control Test	0.080	23:23
Air Blank	0.000	23:24
Diagnostics Check	OK	23:24

Data Download Date/Time: 10/04/2017 19:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BLANCHARD,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 10/03/2017 03:15:39

Diagnostics Check		OK	03:05
Air Blank		0.000	03:06
Control Test		0.081	03:06
Air Blank		0.000	03:07
Subject Sample	# 1	0.172	03:09
Breath Volume		1.257	03:09
Air Blank		0.000	03:09
Air Blank		0.000	03:11
Subject Sample	# 2	0.180	03:13
Breath Volume		1.617	03:13
Air Blank		0.000	03:14
Control Test		0.081	03:15
Air Blank		0.000	03:15
Diagnostics Check		OK	03:15

Data Download Date/Time: 10/04/2017 19:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000809

INSTRUMENT REGISTERED TO: DIXIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FUTCH,JAMES,M
BREATH TEST OPERATOR AGENCY: FFWCC (Lake City)
ARRESTING OFFICER: FUTCH,JAMES
ARRESTING OFFICER AGENCY: FFWCC (Lake City)
LAST AGENCY INSPECTION: 09/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: SPENCER

DATE/TIME OF TEST: 10/08/2017 04:27:38

Diagnostics Check	OK	04:20
Air Blank	0.000	04:21
Control Test	0.080	04:21
Air Blank	0.000	04:22
Subject Sample # 1	0.171	04:22
Breath Volume	2.093	04:22
Air Blank	0.000	04:23
Air Blank	0.000	04:25
Subject Sample # 2	0.178	04:26
Breath Volume	2.269	04:26
Air Blank	0.000	04:26
Control Test	0.082	04:27
Air Blank	0.000	04:27
Diagnostics Check	OK	04:27

Data Download Date/Time: 10/24/2017 10:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000811

INSTRUMENT REGISTERED TO: PERRY POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CANNON,CHAD,M
BREATH TEST OPERATOR AGENCY: Perry PD
ARRESTING OFFICER: CANNON,CHAD
ARRESTING OFFICER AGENCY: Perry PD
LAST AGENCY INSPECTION: 09/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 390314
DRY GAS EXPIRY DATE: 04/19/2019
VIOLATION CODE: DUI
ARREST TIME: 10:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:10

SUBJECT'S LAST NAME: HUNTER

DATE/TIME OF TEST: 10/12/2017 10:37:55

Diagnostics Check	OK	10:31
Air Blank	0.000	10:32
Control Test	0.077	10:32
Air Blank	0.000	10:33
Subject Sample # 1	0.000	10:33
Breath Volume	1.117	10:33
Air Blank	0.000	10:34
Air Blank	0.000	10:36
Subject Sample # 2	0.000	10:36
Breath Volume	1.222	10:36
Air Blank	0.000	10:37
Control Test	0.077	10:37
Air Blank	0.000	10:37
Diagnostics Check	OK	10:37

Data Download Date/Time: 10/30/2017 16:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000811

INSTRUMENT REGISTERED TO: PERRY POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARNOLD,JOHN,S
BREATH TEST OPERATOR AGENCY: Perry PD
ARRESTING OFFICER: WARREN,BRIAN
ARRESTING OFFICER AGENCY: Taylor County SO
LAST AGENCY INSPECTION: 09/28/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 390314
DRY GAS EXPIRY DATE: 04/19/2019
VIOLATION CODE: DUI
ARREST TIME: 11:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 12:08

SUBJECT'S LAST NAME: FOREHAND

DATE/TIME OF TEST: 10/14/2017 12:36:58

Diagnostics Check	OK	12:30
Air Blank	0.000	12:30
Control Test	0.078	12:30
Air Blank	0.000	12:31
Subject Sample # 1	0.139	12:32
Breath Volume	1.941	12:32
Air Blank	0.000	12:32
Air Blank	0.000	12:34
Subject Sample # 2	0.127	12:35
Breath Volume	2.339	12:35
Air Blank	0.000	12:36
Control Test	0.078	12:36
Air Blank	0.000	12:36
Diagnostics Check	OK	12:36

Data Download Date/Time: 10/30/2017 16:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JULIOS,ANTHONY,J
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: GARTNER,A
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 09/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A2
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: OUELLETTE

DATE/TIME OF TEST: 10/01/2017 03:17:19

Diagnostics Check		OK	03:10
Air Blank		0.000	03:11
Control Test		0.077	03:11
Air Blank		0.000	03:11
Subject Sample	# 1	0.265	03:12
Breath Volume		4.070	03:12
Air Blank		0.000	03:13
Air Blank		0.000	03:14
Subject Sample	# 2	0.269	03:15
Breath Volume		3.824	03:15
Air Blank		0.000	03:16
Control Test		0.076	03:16
Air Blank		0.000	03:17
Diagnostics Check		OK	03:17

Data Download Date/Time: 10/08/2017 19:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JULIOS,ANTHONY,J
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: MATHIS,J
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 09/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A2
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: EVANS

DATE/TIME OF TEST: 10/02/2017 00:14:20

Diagnostics Check	OK	00:05	
Air Blank	0.000	00:06	
Control Test	0.078	00:06	
Air Blank	0.000	00:07	
Subject Sample # 1	0.244	00:08	
Breath Volume	1.285	00:08	
Air Blank	0.000	00:09	
Air Blank	0.000	00:11	
Subject Sample # 2	0.246	00:12	
Breath Volume	1.617	00:12	
Air Blank	0.000	00:13	
Control Test	0.074	00:13	Control Outside Tolerance
Air Blank	0.000	00:14	

Data Download Date/Time: 10/08/2017 19:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JULIOS,ANTHONY,J
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: BROWN,R
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 09/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A2
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:17

SUBJECT'S LAST NAME: HARLOW

DATE/TIME OF TEST: 10/02/2017 00:44:02

Diagnostics Check		OK	00:37
Air Blank		0.000	00:37
Control Test		0.076	00:38
Air Blank		0.000	00:38
Subject Sample	# 1	0.267	00:39
Breath Volume		2.019	00:39
Air Blank		0.000	00:40
Air Blank		0.000	00:41
Subject Sample	# 2	0.278	00:42
Breath Volume		2.437	00:42
Air Blank		0.000	00:43
Control Test		0.077	00:43
Air Blank		0.000	00:43
Diagnostics Check		OK	00:44

Data Download Date/Time: 10/08/2017 19:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JULIOS,ANTHONY,J
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: MATHIS,J
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 09/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A2
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: FALLICA

DATE/TIME OF TEST: 10/05/2017 02:11:50

Diagnostics Check	OK	02:05
Air Blank	0.000	02:06
Control Test	0.077	02:06
Air Blank	0.000	02:07
Subject Sample # 1	0.000	02:07
Breath Volume	2.996	02:07
Air Blank	0.000	02:07
Air Blank	0.000	02:10
Subject Sample # 2	0.000	02:10
Breath Volume	2.859	02:10
Air Blank	0.000	02:10
Control Test	0.077	02:11
Air Blank	0.000	02:11
Diagnostics Check	OK	02:11

Data Download Date/Time: 10/08/2017 19:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COCHRANE,AARON,M
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: COCHRANE,A
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 09/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A2
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:32

SUBJECT'S LAST NAME: CALDWELL

DATE/TIME OF TEST: 10/07/2017 00:04:32

Diagnostics Check	OK	00:01	
Air Blank	0.000	00:01	
Control Test	0.077	00:02	
Air Blank	0.000	00:02	
Subject Sample	# 1	00:03	Subject Test Refused
Breath Volume	0.000	00:03	
Air Blank	0.000	00:03	
Control Test	0.078	00:03	
Air Blank	0.000	00:04	
Diagnostics Check	OK	00:04	

Data Download Date/Time: 10/08/2017 19:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASCITELLI,V,D
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: LAVALLE,D
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 09/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A2
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:47

SUBJECT'S LAST NAME: RAMSEY

DATE/TIME OF TEST: 10/07/2017 03:14:25

Diagnostics Check	OK	03:07
Air Blank	0.000	03:08
Control Test	0.078	03:08
Air Blank	0.000	03:09
Subject Sample # 1	0.169	03:09
Breath Volume	2.152	03:09
Air Blank	0.000	03:10
Air Blank	0.000	03:12
Subject Sample # 2	0.170	03:12
Breath Volume	1.285	03:12
Air Blank	0.000	03:13
Control Test	0.077	03:13
Air Blank	0.000	03:14
Diagnostics Check	OK	03:14

Data Download Date/Time: 10/08/2017 19:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COCHRANE,AARON,M
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: COCHRANE,AARON
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 09/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A2
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:22

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 10/08/2017 01:52:15

Diagnostics Check	OK	01:49	
Air Blank	0.000	01:49	
Control Test	0.078	01:49	
Air Blank	0.000	01:50	
Subject Sample	# 1	01:50	Subject Test Refused
Breath Volume	0.000	01:50	
Air Blank	0.000	01:51	
Control Test	0.078	01:51	
Air Blank	0.000	01:52	
Diagnostics Check	OK	01:52	

Data Download Date/Time: 10/08/2017 19:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MENDOZA,LINSEY,R
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: MENDOZA,LINSEY
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 10/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 437738
DRY GAS EXPIRY DATE: 06/08/2019
VIOLATION CODE: DUI
ARREST TIME: 10:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: HAWK

DATE/TIME OF TEST: 10/24/2017 23:25:18

Diagnostics Check	OK	23:18
Air Blank	0.000	23:19
Control Test	0.080	23:19
Air Blank	0.000	23:20
Subject Sample # 1	0.017	23:20
Breath Volume	2.425	23:20
Air Blank	0.000	23:21
Air Blank	0.000	23:23
Subject Sample # 2	0.013	23:23
Breath Volume	1.558	23:23
Air Blank	0.000	23:24
Control Test	0.079	23:24
Air Blank	0.000	23:25
Diagnostics Check	OK	23:25

Data Download Date/Time: 11/28/2017 18:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SOUZA,MAYKOLL,G
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: SOUZA,MAYKOLL
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 10/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 64476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 23:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:57

SUBJECT'S LAST NAME: SAPCOE

DATE/TIME OF TEST: 10/26/2017 00:24:27

Diagnostics Check	OK	00:17
Air Blank	0.000	00:17
Control Test	0.080	00:18
Air Blank	0.000	00:18
Subject Sample # 1	0.010	00:19
Breath Volume	3.484	00:19
Air Blank	0.000	00:20
Air Blank	0.000	00:22
Subject Sample # 2	0.008	00:23
Breath Volume	1.671	00:23
Air Blank	0.000	00:23
Control Test	0.080	00:23
Air Blank	0.000	00:24
Diagnostics Check	OK	00:24

Data Download Date/Time: 11/28/2017 18:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASTELLANO,BRENT,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: CASTELLANO,BRENT
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 10/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 23:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: DAUGHENBAUGH

DATE/TIME OF TEST: 10/26/2017 00:49:52

Diagnostics Check	OK	00:47	
Air Blank	0.000	00:47	
Control Test	0.081	00:47	
Air Blank	0.000	00:48	
Subject Sample	# 1	00:48	Subject Test Refused
Breath Volume	0.000	00:48	
Air Blank	0.000	00:48	
Control Test	0.081	00:49	
Air Blank	0.000	00:49	
Diagnostics Check	OK	00:49	

Data Download Date/Time: 11/28/2017 19:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SOUZA,MAYKOLL,G
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: SOUZA,MAYKOLL
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 10/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 00:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:26

SUBJECT'S LAST NAME: BOYD JR

DATE/TIME OF TEST: 10/27/2017 01:55:29

Diagnostics Check	OK	01:46
Air Blank	0.000	01:46
Control Test	0.081	01:47
Air Blank	0.000	01:47
Subject Sample # 1	0.174	01:50
Breath Volume	1.464	01:50
Air Blank	0.000	01:50
Air Blank	0.000	01:52
Subject Sample # 2	0.165	01:53
Breath Volume	2.031	01:53
Air Blank	0.000	01:54
Control Test	0.080	01:54
Air Blank	0.000	01:55
Diagnostics Check	OK	01:55

Data Download Date/Time: 11/28/2017 19:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BABCOCK,GORDON,D
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: HERNANDEZ,MANUEL
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 10/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 03:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: BABCOCK

DATE/TIME OF TEST: 10/27/2017 04:41:21

Diagnostics Check	OK	04:34
Air Blank	0.000	04:35
Control Test	0.080	04:35
Air Blank	0.000	04:36
Subject Sample # 1	0.250	04:36
Breath Volume	3.917	04:36
Air Blank	0.000	04:37
Air Blank	0.000	04:39
Subject Sample # 2	0.249	04:39
Breath Volume	4.101	04:39
Air Blank	0.000	04:40
Control Test	0.080	04:40
Air Blank	0.000	04:41
Diagnostics Check	OK	04:41

Data Download Date/Time: 11/28/2017 19:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,GARY,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: MILLER,GARY
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 10/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 23:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:01

SUBJECT'S LAST NAME: MINOR

DATE/TIME OF TEST: 10/29/2017 00:30:02

Diagnostics Check	OK	00:23
Air Blank	0.000	00:23
Control Test	0.080	00:23
Air Blank	0.000	00:24
Subject Sample # 1	0.115	00:25
Breath Volume	1.429	00:25
Air Blank	0.000	00:26
Air Blank	0.000	00:27
Subject Sample # 2	0.119	00:28
Breath Volume	2.289	00:28
Air Blank	0.000	00:29
Control Test	0.080	00:29
Air Blank	0.000	00:29
Diagnostics Check	OK	00:30

Data Download Date/Time: 11/28/2017 19:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACOBSON,JAMES,W
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: RAFFERTY,C
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 09/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A3
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 15:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:10

SUBJECT'S LAST NAME: RILEY

DATE/TIME OF TEST: 10/04/2017 16:36:01

Diagnostics Check	OK	16:30	
Air Blank	0.000	16:31	
Control Test	0.079	16:31	
Air Blank	0.000	16:32	
Subject Sample # 1	0.000	16:32	
Breath Volume	3.199	16:32	
Air Blank	0.000	16:32	
Air Blank	0.000	16:34	
Subject Sample # 2	0.000	16:35	
Breath Volume	3.097	16:35	
Air Blank		16:35	RFI Detect
Air Blank	0.000	16:36	

Data Download Date/Time: 10/26/2017 14:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARLOW,V,
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: RAFFERTY,C
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 09/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A3
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:13

SUBJECT'S LAST NAME: LANGSTON

DATE/TIME OF TEST: 10/05/2017 00:41:08

Diagnostics Check	OK	00:35
Air Blank	0.000	00:35
Control Test	0.079	00:35
Air Blank	0.000	00:36
Subject Sample # 1	0.000	00:36
Breath Volume	2.804	00:36
Air Blank	0.000	00:37
Air Blank	0.000	00:39
Subject Sample # 2	0.000	00:39
Breath Volume	2.277	00:39
Air Blank	0.000	00:40
Control Test	0.077	00:40
Air Blank	0.000	00:41
Diagnostics Check	OK	00:41

Data Download Date/Time: 10/26/2017 14:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARLOW,VINESSA,
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: ROSS,C
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 09/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A3
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:14

SUBJECT'S LAST NAME: CLARK

DATE/TIME OF TEST: 10/05/2017 22:56:36

Diagnostics Check		OK	22:41	
Air Blank		0.000	22:42	
Control Test		0.080	22:42	
Air Blank		0.000	22:43	
Subject Sample	# 1	0.125	22:46	Volume Not Met
Breath Volume		0.933	22:46	
Air Blank		0.000	22:47	
Air Blank		0.000	22:48	
Subject Sample	# 2	0.132	22:51	
Breath Volume		1.179	22:51	
Air Blank		0.000	22:52	
Air Blank		0.000	22:53	
Subject Sample	# 3	0.132	22:54	
Breath Volume		1.109	22:54	
Air Blank		0.000	22:55	
Control Test		0.075	22:55	
Air Blank		0.000	22:56	
Diagnostics Check		OK	22:56	

Data Download Date/Time: 10/26/2017 14:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEWIS,JARROD,W
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: RAFFERTY,CHRIS
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 09/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A3
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 17:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:37

SUBJECT'S LAST NAME: G650519726880

DATE/TIME OF TEST: 10/11/2017 19:09:39

Diagnostics Check	OK	19:02
Air Blank	0.000	19:03
Control Test	0.079	19:03
Air Blank	0.000	19:04
Subject Sample # 1	0.268	19:04
Breath Volume	2.054	19:04
Air Blank	0.000	19:05
Air Blank	0.000	19:07
Subject Sample # 2	0.270	19:07
Breath Volume	1.511	19:07
Air Blank	0.000	19:08
Control Test	0.078	19:09
Air Blank	0.000	19:09
Diagnostics Check	OK	19:09

Data Download Date/Time: 10/26/2017 14:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEWIS,JARROD,W
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: RAFFERTY,CHRIS
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 09/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A3
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:01

SUBJECT'S LAST NAME: FOWLER

DATE/TIME OF TEST: 10/13/2017 01:25:57

Diagnostics Check	OK	01:22	
Air Blank	0.000	01:23	
Control Test	0.080	01:23	
Air Blank	0.000	01:24	
Subject Sample	# 1	01:24	Subject Test Refused
Breath Volume	0.000	01:24	
Air Blank	0.000	01:25	
Control Test	0.078	01:25	
Air Blank	0.000	01:25	
Diagnostics Check	OK	01:25	

Data Download Date/Time: 10/26/2017 14:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACOBSON,JAMES,W
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: BUTLER,E
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 09/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A3
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 10:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:09

SUBJECT'S LAST NAME: GRIFFIS

DATE/TIME OF TEST: 10/14/2017 11:40:28

Diagnostics Check	OK	11:31	
Air Blank	0.000	11:31	
Control Test	0.079	11:32	
Air Blank	0.000	11:32	
Subject Sample	# 1	11:35	No Sample Provided
Breath Volume	0.000	11:35	
Air Blank	0.000	11:36	
Air Blank	0.000	11:38	
Subject Sample	# 2	11:39	Subject Test Refused
Breath Volume	0.000	11:39	
Air Blank	0.000	11:39	
Control Test	0.080	11:39	
Air Blank	0.000	11:40	
Diagnostics Check	OK	11:40	

Data Download Date/Time: 10/26/2017 14:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUTTON,EDWIN,C
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: RAFFERTY,C
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 09/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A3
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:42

SUBJECT'S LAST NAME: BREEDEN

DATE/TIME OF TEST: 10/15/2017 00:09:50

Diagnostics Check	OK	00:03
Air Blank	0.000	00:03
Control Test	0.079	00:04
Air Blank	0.000	00:04
Subject Sample # 1	0.114	00:05
Breath Volume	1.554	00:05
Air Blank	0.000	00:05
Air Blank	0.000	00:07
Subject Sample # 2	0.110	00:08
Breath Volume	1.324	00:08
Air Blank	0.000	00:08
Control Test	0.078	00:09
Air Blank	0.000	00:09
Diagnostics Check	OK	00:09

Data Download Date/Time: 10/26/2017 14:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARLOW,V,
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: TAN,R
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 09/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A3
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: WERLINE

DATE/TIME OF TEST: 10/16/2017 00:06:52

Diagnostics Check	OK	00:00
Air Blank	0.000	00:00
Control Test	0.079	00:01
Air Blank	0.000	00:01
Subject Sample # 1	0.032	00:02
Breath Volume	1.761	00:02
Air Blank	0.000	00:03
Air Blank	0.000	00:04
Subject Sample # 2	0.029	00:05
Breath Volume	1.878	00:05
Air Blank	0.000	00:05
Control Test	0.077	00:06
Air Blank	0.000	00:06
Diagnostics Check	OK	00:06

Data Download Date/Time: 10/26/2017 14:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,TRAVIS,D
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: MCCAIG,ASHLEY
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 09/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A3
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:37

SUBJECT'S LAST NAME: BIRCH

DATE/TIME OF TEST: 10/17/2017 01:06:10

Diagnostics Check	OK	00:57
Air Blank	0.000	00:57
Control Test	0.078	00:58
Air Blank	0.000	00:58
Subject Sample # 1	0.171	01:00
Breath Volume	1.667	01:00
Air Blank	0.000	01:01
Air Blank	0.000	01:03
Subject Sample # 2	0.173	01:04
Breath Volume	1.281	01:04
Air Blank	0.000	01:05
Control Test	0.077	01:05
Air Blank	0.000	01:06
Diagnostics Check	OK	01:06

Data Download Date/Time: 10/26/2017 14:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH, TRAVIS, D
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: TUGGEY, TYLER
ARRESTING OFFICER AGENCY: Wildwood PD
LAST AGENCY INSPECTION: 09/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A3
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 18:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:24

SUBJECT'S LAST NAME: SAUNDERS

DATE/TIME OF TEST: 10/22/2017 20:25:27

Diagnostics Check	OK	20:20	
Air Blank	0.000	20:20	
Control Test	0.079	20:21	
Air Blank	0.000	20:21	
Subject Sample	# 1	20:24	Subject Test Refused
Breath Volume	0.000	20:24	
Air Blank	0.000	20:24	
Control Test	0.078	20:24	
Air Blank	0.000	20:25	
Diagnostics Check	OK	20:25	

Data Download Date/Time: 10/26/2017 14:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARSHALL,JARROD,C
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: PETTIT,R
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 10/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A3
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:01

SUBJECT'S LAST NAME: BRYANT

DATE/TIME OF TEST: 10/27/2017 03:35:57

Diagnostics Check	OK	03:28
Air Blank	0.000	03:29
Control Test	0.080	03:29
Air Blank	0.000	03:30
Subject Sample # 1	0.211	03:30
Breath Volume	1.234	03:30
Air Blank	0.000	03:31
Air Blank	0.000	03:33
Subject Sample # 2	0.213	03:34
Breath Volume	2.085	03:34
Air Blank	0.000	03:34
Control Test	0.080	03:35
Air Blank	0.000	03:35
Diagnostics Check	OK	03:35

Data Download Date/Time: 11/12/2017 13:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARLOW,VINESSA,
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: SEARS,T
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 10/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A3
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:12

SUBJECT'S LAST NAME: HENSLEY

DATE/TIME OF TEST: 10/28/2017 20:38:29

Diagnostics Check	OK	20:35	
Air Blank	0.000	20:36	
Control Test	0.079	20:36	
Air Blank	0.000	20:36	
Subject Sample	# 1	20:37	Subject Test Refused
Breath Volume	0.000	20:37	
Air Blank	0.000	20:37	
Control Test	0.079	20:37	
Air Blank	0.000	20:38	
Diagnostics Check	OK	20:38	

Data Download Date/Time: 11/12/2017 13:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARLOW,VINESSA,
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: ROSS,C
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 10/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A3
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:33

SUBJECT'S LAST NAME: ABREU

DATE/TIME OF TEST: 10/29/2017 01:12:38

Diagnostics Check	OK	01:05
Air Blank	0.000	01:05
Control Test	0.079	01:05
Air Blank	0.000	01:06
Subject Sample # 1	0.119	01:07
Breath Volume	1.773	01:07
Air Blank	0.000	01:08
Air Blank	0.000	01:10
Subject Sample # 2	0.121	01:10
Breath Volume	2.125	01:10
Air Blank	0.000	01:11
Control Test	0.077	01:11
Air Blank	0.000	01:12
Diagnostics Check	OK	01:12

Data Download Date/Time: 11/12/2017 13:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARSHALL,JARROD,C
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: ORTIZ,B
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 10/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A3
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:58

SUBJECT'S LAST NAME: VOZNIAK

DATE/TIME OF TEST: 10/31/2017 05:26:16

Diagnostics Check	OK	05:20
Air Blank	0.000	05:20
Control Test	0.079	05:20
Air Blank	0.000	05:21
Subject Sample # 1	0.000	05:22
Breath Volume	3.269	05:22
Air Blank	0.000	05:22
Air Blank	0.000	05:24
Subject Sample # 2	0.000	05:24
Breath Volume	2.535	05:24
Air Blank	0.000	05:25
Control Test	0.079	05:25
Air Blank	0.000	05:26
Diagnostics Check	OK	05:26

Data Download Date/Time: 11/12/2017 13:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHOETTNER,ADRIAN,J
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: TITUS,J
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 10/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:58

SUBJECT'S LAST NAME: CASTINGS

DATE/TIME OF TEST: 10/19/2017 04:40:56

Diagnostics Check		OK	04:33
Air Blank		0.000	04:33
Control Test		0.078	04:34
Air Blank		0.000	04:34
Subject Sample	# 1	0.195	04:35
Breath Volume		2.601	04:35
Air Blank		0.000	04:36
Air Blank		0.000	04:37
Subject Sample	# 2	0.193	04:39
Breath Volume		1.972	04:39
Air Blank		0.000	04:39
Control Test		0.078	04:40
Air Blank		0.000	04:40
Diagnostics Check		OK	04:40

Data Download Date/Time: 11/15/2017 10:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLOYD,DANIEL,Z
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: HOLTZHOUSE,DEP
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 10/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:23

SUBJECT'S LAST NAME: GRANGER

DATE/TIME OF TEST: 10/20/2017 23:58:49

Diagnostics Check	OK	23:55	
Air Blank	0.000	23:55	
Control Test	0.078	23:55	
Air Blank	0.000	23:56	
Subject Sample	# 1	23:58	Sequence Aborted
Breath Volume	0.000	23:58	
Air Blank	0.000	23:58	

Data Download Date/Time: 11/15/2017 10:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLOYD,DANIEL,Z
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: HOLTZHOUSE,DEP
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 10/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:59

SUBJECT'S LAST NAME: GRANGER

DATE/TIME OF TEST: 10/21/2017 00:28:51

Diagnostics Check	OK	00:20
Air Blank	0.000	00:20
Control Test	0.078	00:20
Air Blank	0.000	00:21
Subject Sample # 1	0.205	00:23
Breath Volume	1.511	00:23
Air Blank	0.000	00:24
Air Blank	0.000	00:26
Subject Sample # 2	0.214	00:27
Breath Volume	1.320	00:27
Air Blank	0.000	00:27
Control Test	0.076	00:28
Air Blank	0.000	00:28
Diagnostics Check	OK	00:28

Data Download Date/Time: 11/15/2017 10:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHOETTNER,ADRIAN,J
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: KJEKSTAD,E
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 10/18/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 18:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:27

SUBJECT'S LAST NAME: EILAND

DATE/TIME OF TEST: 10/24/2017 19:53:07

Diagnostics Check	OK	19:50	
Air Blank	0.000	19:50	
Control Test	0.079	19:50	
Air Blank	0.000	19:51	
Subject Sample	# 1	19:51	Subject Test Refused
Breath Volume	0.000	19:51	
Air Blank	0.000	19:52	
Control Test	0.078	19:52	
Air Blank	0.000	19:53	
Diagnostics Check	OK	19:53	

Data Download Date/Time: 11/15/2017 10:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LATHAN,MICHAEL,S
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: FERRER,F
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 10/18/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:08

SUBJECT'S LAST NAME: MUNOZ

DATE/TIME OF TEST: 10/26/2017 23:43:59

Diagnostics Check	OK	23:29	
Air Blank	0.000	23:29	
Control Test	0.078	23:29	
Air Blank	0.000	23:30	
Subject Sample	# 1 0.118	23:31	
Breath Volume	2.085	23:31	
Air Blank	0.000	23:31	
Air Blank	0.000	23:33	
Subject Sample	# 2 0.108	23:36	Volume Not Met
Breath Volume	0.125	23:36	
Air Blank	0.000	23:37	
Air Blank	0.000	23:39	
Subject Sample	# 3 0.126	23:42	Volume Not Met
Breath Volume	0.070	23:42	
Air Blank	0.000	23:43	
Control Test	0.077	23:43	
Air Blank	0.000	23:43	
Diagnostics Check	OK	23:43	

Data Download Date/Time: 11/15/2017 10:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLOYD,DANIEL,Z
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: BANG,DEP
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 10/18/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:26

SUBJECT'S LAST NAME: RUTHERFORD

DATE/TIME OF TEST: 10/31/2017 23:58:14

Diagnostics Check	OK	23:46	
Air Blank	0.000	23:47	
Control Test	0.079	23:47	
Air Blank	0.000	23:48	
Subject Sample	# 1	23:51	No Sample Provided
Breath Volume	0.000	23:51	
Air Blank	0.000	23:51	
Air Blank	0.000	23:53	
Subject Sample	# 2	23:56	No Sample Provided
Breath Volume	0.000	23:56	
Air Blank	0.000	23:57	
Control Test	0.079	23:57	
Air Blank	0.000	23:58	
Diagnostics Check	OK	23:58	

Data Download Date/Time: 11/15/2017 10:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,CRYSTAL,L
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: DOT,DOT
ARRESTING OFFICER AGENCY: CCA Citrus
LAST AGENCY INSPECTION: 09/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: ZERO TOL
ARREST TIME: 10:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:09

SUBJECT'S LAST NAME: BAIZE

DATE/TIME OF TEST: 10/02/2017 10:36:12

Diagnostics Check	OK	10:30
Air Blank	0.000	10:30
Control Test	0.080	10:30
Air Blank	0.000	10:31
Subject Sample # 1	0.000	10:31
Breath Volume	1.359	10:31
Air Blank	0.000	10:32
Air Blank	0.000	10:34
Subject Sample # 2	0.000	10:34
Breath Volume	1.550	10:34
Air Blank	0.000	10:35
Control Test	0.081	10:35
Air Blank	0.000	10:36
Diagnostics Check	OK	10:36

Data Download Date/Time: 10/18/2017 10:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,CRYSTAL,L
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: DOT,DOT
ARRESTING OFFICER AGENCY: CCA Citrus
LAST AGENCY INSPECTION: 09/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: ZERO TOL
ARREST TIME: 10:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:09

SUBJECT'S LAST NAME: BAIZE

DATE/TIME OF TEST: 10/02/2017 10:44:18

Diagnostics Check	OK	10:38
Air Blank	0.000	10:38
Control Test	0.081	10:38
Air Blank	0.000	10:39
Subject Sample # 1	0.000	10:39
Breath Volume	1.253	10:39
Air Blank	0.000	10:40
Air Blank	0.000	10:42
Subject Sample # 2	0.000	10:42
Breath Volume	1.609	10:42
Air Blank	0.000	10:43
Control Test	0.081	10:43
Air Blank	0.000	10:44
Diagnostics Check	OK	10:44

Data Download Date/Time: 10/18/2017 10:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LATHAN,MICHAEL,S
BREATH TEST OPERATOR AGENCY: Citrus County SO
ARRESTING OFFICER: TAYLOR,T
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 09/13/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:05

SUBJECT'S LAST NAME: COOPER

DATE/TIME OF TEST: 10/03/2017 22:43:26

Diagnostics Check	OK	22:40	
Air Blank	0.000	22:40	
Control Test	0.080	22:41	
Air Blank	0.000	22:41	
Subject Sample	# 1	22:42	Subject Test Refused
Breath Volume	0.000	22:42	
Air Blank	0.000	22:42	
Control Test	0.079	22:42	
Air Blank	0.000	22:43	
Diagnostics Check	OK	22:43	

Data Download Date/Time: 10/18/2017 10:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHOETTNER,ADRIAN,J
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: BLOTZ,G
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 09/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: GIBSON

DATE/TIME OF TEST: 10/05/2017 04:04:45

Diagnostics Check	OK	04:01	
Air Blank	0.000	04:02	
Control Test	0.080	04:02	
Air Blank	0.000	04:03	
Subject Sample	# 1	04:03	Subject Test Refused
Breath Volume	0.000	04:03	
Air Blank	0.000	04:03	
Control Test	0.080	04:04	
Air Blank	0.000	04:04	
Diagnostics Check	OK	04:04	

Data Download Date/Time: 10/18/2017 10:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLOYD,DANIEL,Z
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: HOLTZHOUSE,DEP
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 09/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 18:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:52

SUBJECT'S LAST NAME: PARKER

DATE/TIME OF TEST: 10/08/2017 19:20:02

Diagnostics Check	OK	19:13
Air Blank	0.000	19:13
Control Test	0.080	19:13
Air Blank	0.000	19:14
Subject Sample # 1	0.221	19:14
Breath Volume	1.195	19:14
Air Blank	0.000	19:15
Air Blank	0.000	19:17
Subject Sample # 2	0.219	19:18
Breath Volume	1.187	19:18
Air Blank	0.000	19:19
Control Test	0.079	19:19
Air Blank	0.000	19:19
Diagnostics Check	OK	19:20

Data Download Date/Time: 10/18/2017 10:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LATHAN,MICHAEL,S
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: TAYLOR,T
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 09/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: MORTON

DATE/TIME OF TEST: 10/09/2017 02:24:25

Diagnostics Check	OK	02:20	
Air Blank	0.000	02:20	
Control Test	0.079	02:21	
Air Blank	0.000	02:21	
Subject Sample	# 1	02:23	Subject Test Refused
Breath Volume	0.000	02:23	
Air Blank	0.000	02:23	
Control Test	0.081	02:23	
Air Blank	0.000	02:24	
Diagnostics Check	OK	02:24	

Data Download Date/Time: 10/18/2017 10:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHOETTNER,ADRIAN,J
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: CLOUD,T.
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 09/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 18:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:41

SUBJECT'S LAST NAME: BOXWELL

DATE/TIME OF TEST: 10/14/2017 19:21:12

Diagnostics Check	OK	19:14
Air Blank	0.000	19:14
Control Test	0.080	19:15
Air Blank	0.000	19:15
Subject Sample # 1	0.297	19:16
Breath Volume	1.761	19:16
Air Blank	0.000	19:17
Air Blank	0.000	19:18
Subject Sample # 2	0.306	19:19
Breath Volume	1.210	19:19
Air Blank	0.000	19:20
Control Test	0.080	19:20
Air Blank	0.000	19:21
Diagnostics Check	OK	19:21

Data Download Date/Time: 10/18/2017 10:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLOYD,DANIEL,Z
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: HOLTZHOUSE,DEP
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 10/18/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:23

SUBJECT'S LAST NAME: GRANGER

DATE/TIME OF TEST: 10/20/2017 23:50:59

Diagnostics Check	OK	23:43	
Air Blank	0.000	23:43	
Control Test	0.080	23:44	
Air Blank	0.000	23:44	
Subject Sample	# 1 0.212	23:47	Volume Not Met
Breath Volume	1.050	23:47	
Air Blank	0.000	23:48	
Air Blank	0.000	23:50	
Subject Sample	# 2	23:50	RFI Detect
Breath Volume	0.000	23:50	
Air Blank		23:50	RFI Detect
Air Blank		23:50	RFI Detect

Data Download Date/Time: 10/26/2017 14:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLTZHOUSE,JONATHAN,D
BREATH TEST OPERATOR AGENCY: Citrus County SO
ARRESTING OFFICER: HOLTZHOUSE,JONATHAN
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 10/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 09:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:31

SUBJECT'S LAST NAME: FISCHER

DATE/TIME OF TEST: 10/26/2017 09:58:36

Diagnostics Check	OK	09:53	
Air Blank	0.000	09:53	
Control Test	0.081	09:54	
Air Blank	0.000	09:54	
Subject Sample # 1	0.000	09:55	
Breath Volume	3.269	09:55	
Air Blank	0.000	09:55	
Air Blank		09:57	Improper Sample
Air Blank	0.000	09:57	
Control Test	0.082	09:57	
Air Blank	0.000	09:58	
Diagnostics Check	OK	09:58	

Data Download Date/Time: 10/26/2017 14:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORMAN,GERALD,E
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: CORMAN,GERALD
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 09/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: ABSTON

DATE/TIME OF TEST: 10/03/2017 00:52:07

Diagnostics Check	OK	00:48	
Air Blank	0.000	00:49	
Control Test	0.080	00:49	
Air Blank	0.000	00:50	
Subject Sample	# 1	00:50	Subject Test Refused
Breath Volume	0.000	00:50	
Air Blank	0.000	00:51	
Control Test	0.080	00:51	
Air Blank	0.000	00:52	
Diagnostics Check	OK	00:52	

Data Download Date/Time: 10/25/2017 20:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORMAN,GERALD,E
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: CORMAN,GERALD
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 09/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:42

SUBJECT'S LAST NAME: DELGADO MARTINEZ

DATE/TIME OF TEST: 10/08/2017 03:04:42

Diagnostics Check	OK	03:04	
Air Blank		03:04	RFI Detect
Air Blank	0.000	03:04	

Data Download Date/Time: 10/25/2017 20:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORMAN,GERALD,E
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: CORMAN,GERALD
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 09/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:42

SUBJECT'S LAST NAME: DELGADO MARTINEZ

DATE/TIME OF TEST: 10/08/2017 03:13:13

Diagnostics Check	OK	03:06
Air Blank	0.000	03:06
Control Test	0.080	03:07
Air Blank	0.000	03:07
Subject Sample # 1	0.148	03:08
Breath Volume	1.937	03:08
Air Blank	0.000	03:09
Air Blank	0.000	03:11
Subject Sample # 2	0.138	03:11
Breath Volume	2.500	03:11
Air Blank	0.000	03:12
Control Test	0.080	03:12
Air Blank	0.000	03:13
Diagnostics Check	OK	03:13

Data Download Date/Time: 10/25/2017 20:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORMAN,GERALD,E
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: CORMAN,GERALD
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 09/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:04

SUBJECT'S LAST NAME: SHIMMONS

DATE/TIME OF TEST: 10/09/2017 00:30:34

Diagnostics Check	OK	00:24
Air Blank	0.000	00:24
Control Test	0.080	00:25
Air Blank	0.000	00:25
Subject Sample # 1	0.000	00:26
Breath Volume	3.406	00:26
Air Blank	0.000	00:26
Air Blank	0.000	00:28
Subject Sample # 2	0.000	00:29
Breath Volume	3.292	00:29
Air Blank	0.000	00:29
Control Test	0.081	00:29
Air Blank	0.000	00:30
Diagnostics Check	OK	00:30

Data Download Date/Time: 10/25/2017 20:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHESSHER,SANDRA,W
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: CROW,S
ARRESTING OFFICER AGENCY: Umatilla PD
LAST AGENCY INSPECTION: 09/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:43

SUBJECT'S LAST NAME: OWSLEY

DATE/TIME OF TEST: 10/19/2017 22:10:34

Diagnostics Check	OK	22:04
Air Blank	0.000	22:04
Control Test	0.080	22:04
Air Blank	0.000	22:05
Subject Sample # 1	0.126	22:06
Breath Volume	3.218	22:06
Air Blank	0.000	22:06
Air Blank	0.000	22:08
Subject Sample # 2	0.132	22:09
Breath Volume	3.628	22:09
Air Blank	0.000	22:09
Control Test	0.083	22:09
Air Blank	0.000	22:10
Diagnostics Check	OK	22:10

Data Download Date/Time: 10/25/2017 20:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PITCHER,DARRYL,P
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: PITCHER,DARRYL
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 09/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: HAMIL

DATE/TIME OF TEST: 10/19/2017 23:56:17

Diagnostics Check	OK	23:41	
Air Blank	0.000	23:41	
Control Test	0.081	23:41	
Air Blank	0.000	23:42	
Subject Sample	# 1 0.143	23:43	
Breath Volume	2.500	23:43	
Air Blank	0.000	23:44	
Air Blank	0.000	23:45	
Subject Sample	# 2 0.121	23:49	Volume Not Met
Breath Volume	0.765	23:49	
Air Blank	0.000	23:49	
Air Blank	0.000	23:51	
Subject Sample	# 3 0.136	23:54	Volume Not Met
Breath Volume	0.421	23:54	
Air Blank	0.000	23:55	
Control Test	0.081	23:55	
Air Blank	0.000	23:56	
Diagnostics Check	OK	23:56	

Data Download Date/Time: 10/25/2017 20:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORMAN,GERALD,E
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: CORMAN,GERALD
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 09/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:43

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 10/22/2017 01:09:43

Diagnostics Check	OK	01:06	
Air Blank	0.000	01:07	
Control Test	0.081	01:07	
Air Blank	0.000	01:08	
Subject Sample	# 1	01:08	Subject Test Refused
Breath Volume	0.000	01:08	
Air Blank	0.000	01:08	
Control Test	0.080	01:09	
Air Blank	0.000	01:09	
Diagnostics Check	OK	01:09	

Data Download Date/Time: 10/25/2017 20:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,ETHAN,C
BREATH TEST OPERATOR AGENCY: Clermont PD
ARRESTING OFFICER: BELL,ETHAN
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 09/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 06:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:45

SUBJECT'S LAST NAME: PENA

DATE/TIME OF TEST: 10/24/2017 08:12:11

Diagnostics Check		OK	08:06
Air Blank		0.000	08:06
Control Test		0.079	08:06
Air Blank		0.000	08:07
Subject Sample	# 1	0.228	08:07
Breath Volume		2.050	08:07
Air Blank		0.000	08:08
Air Blank		0.000	08:10
Subject Sample	# 2	0.230	08:10
Breath Volume		1.925	08:10
Air Blank		0.000	08:11
Control Test		0.080	08:11
Air Blank		0.000	08:12
Diagnostics Check		OK	08:12

Data Download Date/Time: 10/25/2017 20:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PITCHER,DARRYL,P
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: PITCHER,DARRYL
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 10/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 10/27/2017 02:01:42

Diagnostics Check	OK	01:50	
Air Blank	0.000	01:50	
Control Test	0.080	01:51	
Air Blank	0.000	01:51	
Subject Sample	# 1	01:54	No Sample Provided
Breath Volume	0.000	01:54	
Air Blank	0.000	01:55	
Air Blank	0.000	01:57	
Subject Sample	# 2	02:00	No Sample Provided
Breath Volume	0.000	02:00	
Air Blank	0.000	02:00	
Control Test	0.081	02:01	
Air Blank	0.000	02:01	
Diagnostics Check	OK	02:01	

Data Download Date/Time: 11/03/2017 21:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLESSING,JARED,A
BREATH TEST OPERATOR AGENCY: Tavares PD
ARRESTING OFFICER: BLESSING,JARED
ARRESTING OFFICER AGENCY: Tavares PD
LAST AGENCY INSPECTION: 10/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: ALBRIGHT

DATE/TIME OF TEST: 10/27/2017 21:15:00

Diagnostics Check	OK	21:12	
Air Blank	0.000	21:12	
Control Test	0.080	21:12	
Air Blank	0.000	21:13	
Subject Sample	# 1	21:13	Subject Test Refused
Breath Volume	0.000	21:13	
Air Blank	0.000	21:14	
Control Test	0.080	21:14	
Air Blank	0.000	21:14	
Diagnostics Check	OK	21:15	

Data Download Date/Time: 11/03/2017 21:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHESSHER,SANDRA,W
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: CHESSHER,SANDRA
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 10/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:47

SUBJECT'S LAST NAME: EDWARDS

DATE/TIME OF TEST: 10/28/2017 04:14:20

Diagnostics Check		OK	04:07
Air Blank		0.000	04:08
Control Test		0.080	04:08
Air Blank		0.000	04:08
Subject Sample	# 1	0.206	04:09
Breath Volume		2.898	04:09
Air Blank		0.000	04:10
Air Blank		0.000	04:12
Subject Sample	# 2	0.211	04:12
Breath Volume		2.437	04:12
Air Blank		0.000	04:13
Control Test		0.080	04:13
Air Blank		0.000	04:14
Diagnostics Check		OK	04:14

Data Download Date/Time: 11/03/2017 21:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PITCHER,DARRYL,P
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: PITCHER,DARRYL
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 10/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 18:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:00

SUBJECT'S LAST NAME: BOUGHTON

DATE/TIME OF TEST: 10/28/2017 19:28:05

Diagnostics Check	OK	19:21
Air Blank	0.000	19:22
Control Test	0.080	19:22
Air Blank	0.000	19:23
Subject Sample # 1	0.000	19:23
Breath Volume	1.113	19:23
Air Blank	0.000	19:24
Air Blank	0.000	19:26
Subject Sample # 2	0.000	19:26
Breath Volume	2.109	19:26
Air Blank	0.000	19:27
Control Test	0.080	19:27
Air Blank	0.000	19:28
Diagnostics Check	OK	19:28

Data Download Date/Time: 11/03/2017 21:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,ETHAN,C
BREATH TEST OPERATOR AGENCY: Clermont PD
ARRESTING OFFICER: BELL,ETHAN
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 09/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1766638
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: HAGAN

DATE/TIME OF TEST: 10/14/2017 02:35:03

Diagnostics Check		OK	02:25
Air Blank		0.000	02:26
Control Test		0.078	02:26
Air Blank		0.000	02:27
Subject Sample	# 1	0.245	02:27
Breath Volume		2.417	02:27
Air Blank		0.000	02:28
Air Blank		0.000	02:29
Subject Sample	# 2	0.218	02:30
Breath Volume		2.710	02:30
Air Blank		0.000	02:31
Air Blank		0.000	02:32
Subject Sample	# 3	0.226	02:33
Breath Volume		1.886	02:33
Air Blank		0.000	02:34
Control Test		0.078	02:34
Air Blank		0.000	02:34
Diagnostics Check		OK	02:35

Data Download Date/Time: 10/25/2017 20:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,ETHAN,C
BREATH TEST OPERATOR AGENCY: Clermont PD
ARRESTING OFFICER: BELL,ETHAN
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 09/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1766638
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:33

SUBJECT'S LAST NAME: KOOPMAN

DATE/TIME OF TEST: 10/15/2017 05:01:47

Diagnostics Check	OK	04:53	
Air Blank	0.000	04:53	
Control Test	0.079	04:53	
Air Blank	0.000	04:54	
Subject Sample	# 1	04:57	No Sample Provided
Breath Volume	0.000	04:57	
Air Blank	0.000	04:58	
Air Blank	0.000	05:00	
Subject Sample	# 2	05:00	Subject Test Refused
Breath Volume	0.000	05:00	
Air Blank	0.000	05:00	
Control Test	0.078	05:01	
Air Blank	0.000	05:01	
Diagnostics Check	OK	05:01	

Data Download Date/Time: 10/25/2017 20:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,ETHAN,C
BREATH TEST OPERATOR AGENCY: Clermont PD
ARRESTING OFFICER: CHRISTIE,KEVIN
ARRESTING OFFICER AGENCY: Groveland PD
LAST AGENCY INSPECTION: 09/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1766638
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:47

SUBJECT'S LAST NAME: REALPE

DATE/TIME OF TEST: 10/16/2017 05:13:59

Diagnostics Check		OK	05:07
Air Blank		0.000	05:08
Control Test		0.079	05:08
Air Blank		0.000	05:09
Subject Sample	# 1	0.150	05:09
Breath Volume		2.140	05:09
Air Blank		0.000	05:10
Air Blank		0.000	05:11
Subject Sample	# 2	0.154	05:12
Breath Volume		2.191	05:12
Air Blank		0.000	05:13
Control Test		0.077	05:13
Air Blank		0.000	05:13
Diagnostics Check		OK	05:13

Data Download Date/Time: 10/25/2017 20:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUNLAP,JASON,N
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: DUNLAP,JASON
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 09/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1766638
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: ROOPNARINE

DATE/TIME OF TEST: 10/21/2017 04:34:37

Diagnostics Check	OK	04:22	
Air Blank	0.000	04:23	
Control Test	0.078	04:23	
Air Blank	0.000	04:24	
Subject Sample	# 1	04:27	No Sample Provided
Breath Volume	0.000	04:27	
Air Blank	0.000	04:27	
Air Blank	0.000	04:29	
Subject Sample	# 2	04:32	No Sample Provided
Breath Volume	0.000	04:32	
Air Blank	0.000	04:33	
Control Test	0.079	04:33	
Air Blank	0.000	04:34	
Diagnostics Check	OK	04:34	

Data Download Date/Time: 10/25/2017 20:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUNLAP,JASON,N
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: DUNLAP,JASON
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 09/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1766638
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: ROOPNARINE

DATE/TIME OF TEST: 10/21/2017 04:48:15

Diagnostics Check		OK	04:36	
Air Blank		0.000	04:36	
Control Test		0.077	04:37	
Air Blank		0.000	04:37	
Subject Sample	# 1	0.192	04:40	Volume Not Met
Breath Volume		0.062	04:40	
Air Blank		0.000	04:41	
Air Blank		0.000	04:43	
Subject Sample	# 2	0.172	04:46	Volume Not Met
Breath Volume		0.140	04:46	
Air Blank		0.000	04:47	
Control Test		0.081	04:47	
Air Blank		0.000	04:48	
Diagnostics Check		OK	04:48	

Data Download Date/Time: 10/25/2017 20:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUNLAP,JASON,N
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: DUNLAP,JASON
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 09/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1766638
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: ROOPNARINE

DATE/TIME OF TEST: 10/21/2017 05:00:43

Diagnostics Check		OK	04:50	
Air Blank		0.000	04:50	
Control Test		0.078	04:50	
Air Blank		0.000	04:51	
Subject Sample	# 1	0.177	04:54	Volume Not Met
Breath Volume		0.320	04:54	
Air Blank		0.000	04:55	
Air Blank		0.000	04:57	
Subject Sample	# 2		04:58	Subject Test Refused
Breath Volume		0.304	04:58	
Air Blank		0.000	04:59	
Control Test		0.078	05:00	
Air Blank		0.000	05:00	
Diagnostics Check		OK	05:00	

Data Download Date/Time: 10/25/2017 20:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,ETHAN,C
BREATH TEST OPERATOR AGENCY: Clermont PD
ARRESTING OFFICER: BELL,ETHAN
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 09/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1766638
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 15:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:18

SUBJECT'S LAST NAME: SCHMIDT

DATE/TIME OF TEST: 10/25/2017 16:44:41

Diagnostics Check	OK	16:32	
Air Blank	0.000	16:32	
Control Test	0.078	16:33	
Air Blank	0.000	16:33	
Subject Sample	# 1	16:36	No Sample Provided
Breath Volume	0.000	16:36	
Air Blank	0.000	16:37	
Air Blank	0.000	16:39	
Subject Sample	# 2	0.323	16:39
Breath Volume	1.648	16:39	
Air Blank	0.000	16:40	
Air Blank	0.000	16:42	
Subject Sample	# 3	0.335	16:42
Breath Volume	1.308	16:42	
Air Blank	0.000	16:43	
Control Test	0.076	16:44	
Air Blank	0.000	16:44	
Diagnostics Check	OK	16:44	

Data Download Date/Time: 10/25/2017 20:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FINDLEY,CARLTONJANE,P
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: FINDLEY,CARLTONJANE
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 10/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1766638
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 06:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:32

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 10/26/2017 07:02:51

Diagnostics Check	OK	06:53
Air Blank	0.000	06:54
Control Test	0.079	06:54
Air Blank	0.000	06:55
Subject Sample # 1	0.160	06:57
Breath Volume	1.164	06:57
Air Blank	0.000	06:58
Air Blank	0.000	06:59
Subject Sample # 2	0.158	07:01
Breath Volume	1.589	07:01
Air Blank	0.000	07:01
Control Test	0.078	07:02
Air Blank	0.000	07:02
Diagnostics Check	OK	07:02

Data Download Date/Time: 11/03/2017 21:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEVENS,JEREMY,L
BREATH TEST OPERATOR AGENCY: Leesburg PD
ARRESTING OFFICER: STEVENS,JEREMY
ARRESTING OFFICER AGENCY: Leesburg PD
LAST AGENCY INSPECTION: 09/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 18:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:00

SUBJECT'S LAST NAME: ADAMS

DATE/TIME OF TEST: 10/08/2017 19:35:34

Diagnostics Check	OK	19:31	
Air Blank	0.000	19:32	
Control Test	0.080	19:32	
Air Blank	0.000	19:33	
Subject Sample	# 1	19:34	Interferent Detect
Breath Volume	1.792	19:34	
Air Blank	0.000	19:34	
Control Test	0.076	19:35	Interferent Detect
Air Blank	0.000	19:35	

Data Download Date/Time: 10/28/2017 17:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEVENS, JEREMY, L
BREATH TEST OPERATOR AGENCY: Leesburg PD
ARRESTING OFFICER: STEVENS, JEREMY
ARRESTING OFFICER AGENCY: Leesburg PD
LAST AGENCY INSPECTION: 09/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 18:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:00

SUBJECT'S LAST NAME: A352890861840

DATE/TIME OF TEST: 10/08/2017 19:47:43

Diagnostics Check	OK	19:41
Air Blank	0.000	19:42
Control Test	0.079	19:42
Air Blank	0.000	19:42
Subject Sample # 1	0.147	19:43
Breath Volume	1.531	19:43
Air Blank	0.000	19:43
Air Blank	0.000	19:45
Subject Sample # 2	0.132	19:46
Breath Volume	2.367	19:46
Air Blank	0.000	19:46
Control Test	0.079	19:47
Air Blank	0.000	19:47
Diagnostics Check	OK	19:47

Data Download Date/Time: 10/28/2017 17:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEVENS, JEREMY, L
BREATH TEST OPERATOR AGENCY: Leesburg PD
ARRESTING OFFICER: STEVENS, JEREMY
ARRESTING OFFICER AGENCY: Leesburg PD
LAST AGENCY INSPECTION: 09/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:24

SUBJECT'S LAST NAME: BAUMGARTEN

DATE/TIME OF TEST: 10/11/2017 21:59:46

Diagnostics Check		OK	21:44	
Air Blank		0.000	21:45	
Control Test		0.079	21:45	
Air Blank		0.000	21:46	
Subject Sample	# 1	0.197	21:47	
Breath Volume		1.335	21:47	
Air Blank		0.000	21:47	
Air Blank		0.000	21:49	
Subject Sample	# 2	0.176	21:52	Volume Not Met
Breath Volume		0.941	21:52	
Air Blank		0.000	21:53	
Air Blank		0.000	21:55	
Subject Sample	# 3	0.160	21:58	Volume Not Met
Breath Volume		0.335	21:58	
Air Blank		0.000	21:58	
Control Test		0.076	21:59	
Air Blank		0.000	21:59	
Diagnostics Check		OK	21:59	

Data Download Date/Time: 10/28/2017 17:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEVENS,JEREMY,L
BREATH TEST OPERATOR AGENCY: Leesburg PD
ARRESTING OFFICER: STEVENS,JEREMY
ARRESTING OFFICER AGENCY: Leesburg PD
LAST AGENCY INSPECTION: 09/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:24

SUBJECT'S LAST NAME: BAUMGARTEN

DATE/TIME OF TEST: 10/11/2017 22:09:45

Diagnostics Check	OK	22:02
Air Blank	0.000	22:02
Control Test	0.078	22:03
Air Blank	0.000	22:03
Subject Sample # 1	0.222	22:04
Breath Volume	1.156	22:04
Air Blank	0.000	22:04
Air Blank	0.000	22:06
Subject Sample # 2	0.224	22:08
Breath Volume	1.421	22:08
Air Blank	0.000	22:08
Control Test	0.078	22:09
Air Blank	0.000	22:09
Diagnostics Check	OK	22:09

Data Download Date/Time: 10/28/2017 17:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FOSTER,ANDREW,D
BREATH TEST OPERATOR AGENCY: Leesburg PD
ARRESTING OFFICER: FOSTER,ANDREW
ARRESTING OFFICER AGENCY: Leesburg PD
LAST AGENCY INSPECTION: 09/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: HEYWARD

DATE/TIME OF TEST: 10/16/2017 02:11:15

Diagnostics Check	OK	02:07	
Air Blank	0.000	02:08	
Control Test	0.079	02:08	
Air Blank	0.000	02:09	
Subject Sample	# 1	02:09	Subject Test Refused
Breath Volume	0.000	02:09	
Air Blank	0.000	02:10	
Control Test	0.080	02:10	
Air Blank	0.000	02:11	
Diagnostics Check	OK	02:11	

Data Download Date/Time: 10/28/2017 17:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FOSTER,ANDREW,D
BREATH TEST OPERATOR AGENCY: Leesburg PD
ARRESTING OFFICER: FOSTER,ANDREW
ARRESTING OFFICER AGENCY: Leesburg PD
LAST AGENCY INSPECTION: 09/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: HEYWARD

DATE/TIME OF TEST: 10/16/2017 03:19:12

Diagnostics Check	OK	03:15	
Air Blank	0.000	03:16	
Control Test	0.080	03:16	
Air Blank	0.000	03:17	
Subject Sample	# 1	03:18	Subject Test Refused
Breath Volume	0.058	03:18	
Air Blank		03:18	RFI Detect
Air Blank	0.000	03:19	

Data Download Date/Time: 10/28/2017 17:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000825

INSTRUMENT REGISTERED TO: MOUNT DORA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAYLOR,KEITH,A
BREATH TEST OPERATOR AGENCY: Mount Dora PD
ARRESTING OFFICER: RATLIFF,J
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 09/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 583457
DRY GAS EXPIRY DATE: 10/20/2019
VIOLATION CODE: DUI
ARREST TIME: 14:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:19

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 10/03/2017 15:49:17

Diagnostics Check	OK	15:40
Air Blank	0.000	15:40
Control Test	0.079	15:41
Air Blank	0.000	15:41
Subject Sample # 1	0.152	15:44
Breath Volume	1.558	15:44
Air Blank	0.000	15:45
Air Blank	0.000	15:47
Subject Sample # 2	0.165	15:47
Breath Volume	1.703	15:47
Air Blank	0.000	15:48
Control Test	0.077	15:48
Air Blank	0.000	15:49
Diagnostics Check	OK	15:49

Data Download Date/Time: 10/08/2017 13:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: UMANA,RAUL,J
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: UMANA,RAUL
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 09/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 01:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: BELL

DATE/TIME OF TEST: 10/02/2017 02:51:39

Diagnostics Check	OK	02:45
Air Blank	0.000	02:45
Control Test	0.079	02:46
Air Blank	0.000	02:46
Subject Sample # 1	0.135	02:47
Breath Volume	3.257	02:47
Air Blank	0.000	02:47
Air Blank	0.000	02:49
Subject Sample # 2	0.137	02:50
Breath Volume	2.238	02:50
Air Blank	0.000	02:50
Control Test	0.081	02:51
Air Blank	0.000	02:51
Diagnostics Check	OK	02:51

Data Download Date/Time: 10/24/2017 18:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BALL,CODY,W
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: BALL,CODY
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 09/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 18:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:15

SUBJECT'S LAST NAME: DUNN

DATE/TIME OF TEST: 10/04/2017 19:53:52

Diagnostics Check		OK	19:36	
Air Blank		0.000	19:37	
Control Test		0.079	19:37	
Air Blank		0.000	19:38	
Subject Sample	# 1	0.184	19:41	Volume Not Met
Breath Volume		0.031	19:41	
Air Blank		0.000	19:42	
Air Blank		0.000	19:43	
Subject Sample	# 2	0.213	19:46	
Breath Volume		1.882	19:46	
Air Blank		0.000	19:47	
Air Blank		0.000	19:49	
Subject Sample	# 3	0.159	19:52	Volume Not Met
Breath Volume		0.539	19:52	
Air Blank		0.000	19:52	
Control Test		0.079	19:53	
Air Blank		0.000	19:53	
Diagnostics Check		OK	19:53	

Data Download Date/Time: 10/24/2017 18:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,GARY,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: MILLER,GARY
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 09/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 22:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: HOOPER

DATE/TIME OF TEST: 10/04/2017 23:42:52

Diagnostics Check	OK	23:36
Air Blank	0.000	23:37
Control Test	0.079	23:37
Air Blank	0.000	23:37
Subject Sample # 1	0.076	23:38
Breath Volume	1.714	23:38
Air Blank	0.000	23:39
Air Blank	0.000	23:40
Subject Sample # 2	0.076	23:41
Breath Volume	1.152	23:41
Air Blank	0.000	23:41
Control Test	0.079	23:42
Air Blank	0.000	23:42
Diagnostics Check	OK	23:42

Data Download Date/Time: 10/24/2017 18:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HECKMAN,SHELBY,L
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: HECKMAN,SHELBY
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 09/26/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 23:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:43

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 10/05/2017 00:16:11

Diagnostics Check	OK	00:04	
Air Blank	0.000	00:04	
Control Test	0.079	00:04	
Air Blank	0.000	00:05	
Subject Sample # 1	0.105	00:05	
Breath Volume	1.265	00:05	
Air Blank	0.000	00:06	
Air Blank	0.000	00:08	
Subject Sample # 2	0.094	00:11	Volume Not Met
Breath Volume	0.316	00:11	
Air Blank	0.000	00:11	
Air Blank	0.000	00:13	
Subject Sample # 3	0.106	00:14	
Breath Volume	3.289	00:14	
Air Blank	0.000	00:15	
Control Test	0.077	00:15	
Air Blank	0.000	00:16	
Diagnostics Check	OK	00:16	

Data Download Date/Time: 10/24/2017 18:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,ALYSSA,D
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: WILLIAMS,RONNIE
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 09/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 02:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:11

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 10/06/2017 03:46:17

Diagnostics Check	OK	03:39
Air Blank	0.000	03:39
Control Test	0.080	03:40
Air Blank	0.000	03:40
Subject Sample # 1	0.237	03:41
Breath Volume	3.046	03:41
Air Blank	0.000	03:41
Air Blank	0.000	03:43
Subject Sample # 2	0.235	03:44
Breath Volume	3.878	03:44
Air Blank	0.000	03:45
Control Test	0.081	03:45
Air Blank	0.000	03:46
Diagnostics Check	OK	03:46

Data Download Date/Time: 10/24/2017 18:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,GARY,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: MILLER,GARY
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 09/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 01:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:57

SUBJECT'S LAST NAME: ANDALORO

DATE/TIME OF TEST: 10/07/2017 02:23:44

Diagnostics Check	OK	02:17
Air Blank	0.000	02:17
Control Test	0.080	02:18
Air Blank	0.000	02:18
Subject Sample # 1	0.114	02:19
Breath Volume	1.449	02:19
Air Blank	0.000	02:19
Air Blank	0.000	02:21
Subject Sample # 2	0.116	02:22
Breath Volume	1.132	02:22
Air Blank	0.000	02:22
Control Test	0.079	02:23
Air Blank	0.000	02:23
Diagnostics Check	OK	02:23

Data Download Date/Time: 10/24/2017 18:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SOLLAZZO,C,
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: SOLLAZZO,C
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 09/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 01:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: LANGLEY

DATE/TIME OF TEST: 10/07/2017 02:36:13

Diagnostics Check	OK	02:32	
Air Blank	0.000	02:33	
Control Test	0.080	02:33	
Air Blank	0.000	02:33	
Subject Sample	# 1	02:34	Subject Test Refused
Breath Volume	0.000	02:34	
Air Blank	0.000	02:35	
Control Test	0.080	02:35	
Air Blank	0.000	02:36	
Diagnostics Check	OK	02:36	

Data Download Date/Time: 10/24/2017 18:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SOUZA,MAYKOLL,G
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: SOUZA,MAYKOLL
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 09/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:18

SUBJECT'S LAST NAME: COX

DATE/TIME OF TEST: 10/07/2017 03:51:19

Diagnostics Check		OK	03:38	
Air Blank		0.000	03:39	
Control Test		0.080	03:39	
Air Blank		0.000	03:39	
Subject Sample	# 1	0.169	03:43	Volume Not Met
Breath Volume		0.414	03:43	
Air Blank		0.000	03:43	
Air Blank		0.000	03:45	
Subject Sample	# 2	0.172	03:46	
Breath Volume		1.320	03:46	
Air Blank		0.000	03:46	
Air Blank		0.000	03:48	
Subject Sample	# 3	0.160	03:49	
Breath Volume		1.601	03:49	
Air Blank		0.000	03:50	
Control Test		0.079	03:50	
Air Blank		0.000	03:51	
Diagnostics Check		OK	03:51	

Data Download Date/Time: 10/24/2017 18:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DRAKE, DENISE, A
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: DEAS, JAMES
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 09/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 07:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:12

SUBJECT'S LAST NAME: SAPONARO

DATE/TIME OF TEST: 10/07/2017 07:57:00

Diagnostics Check	OK	07:50
Air Blank	0.000	07:51
Control Test	0.079	07:51
Air Blank	0.000	07:52
Subject Sample # 1	0.143	07:52
Breath Volume	3.203	07:52
Air Blank	0.000	07:53
Air Blank	0.000	07:55
Subject Sample # 2	0.138	07:55
Breath Volume	1.949	07:55
Air Blank	0.000	07:56
Control Test	0.077	07:56
Air Blank	0.000	07:56
Diagnostics Check	OK	07:57

Data Download Date/Time: 10/24/2017 18:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DILLON,TOMAS,R
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: DILLON,TOMAS
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 09/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 11:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: SEEKO

DATE/TIME OF TEST: 10/08/2017 00:52:24

Diagnostics Check	OK	00:45
Air Blank	0.000	00:46
Control Test	0.080	00:46
Air Blank	0.000	00:47
Subject Sample # 1	0.176	00:47
Breath Volume	2.121	00:47
Air Blank	0.000	00:48
Air Blank	0.000	00:50
Subject Sample # 2	0.181	00:50
Breath Volume	2.074	00:50
Air Blank	0.000	00:51
Control Test	0.078	00:51
Air Blank	0.000	00:52
Diagnostics Check	OK	00:52

Data Download Date/Time: 10/24/2017 18:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SOLLAZZO,C,
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: SOLLAZZO,C
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 09/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 22:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:06

SUBJECT'S LAST NAME: GREENE

DATE/TIME OF TEST: 10/08/2017 23:32:11

Diagnostics Check	OK	23:26
Air Blank	0.000	23:26
Control Test	0.080	23:27
Air Blank	0.000	23:27
Subject Sample # 1	0.156	23:27
Breath Volume	2.171	23:27
Air Blank	0.000	23:28
Air Blank	0.000	23:30
Subject Sample # 2	0.153	23:30
Breath Volume	2.121	23:30
Air Blank	0.000	23:31
Control Test	0.079	23:31
Air Blank	0.000	23:32
Diagnostics Check	OK	23:32

Data Download Date/Time: 10/24/2017 18:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WORFEL,KYLE,A
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: WORFEL,KYLE
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 09/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 20:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: GAERTNER

DATE/TIME OF TEST: 10/09/2017 21:23:36

Diagnostics Check	OK	21:16
Air Blank	0.000	21:17
Control Test	0.081	21:17
Air Blank	0.000	21:18
Subject Sample # 1	0.171	21:19
Breath Volume	1.378	21:19
Air Blank	0.000	21:19
Air Blank	0.000	21:21
Subject Sample # 2	0.164	21:22
Breath Volume	1.906	21:22
Air Blank	0.000	21:22
Control Test	0.079	21:23
Air Blank	0.000	21:23
Diagnostics Check	OK	21:23

Data Download Date/Time: 10/24/2017 18:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAKER,MARK,J
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: BAKER,MARK
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 09/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 23:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: ALLISON

DATE/TIME OF TEST: 10/10/2017 00:19:53

Diagnostics Check	OK	00:13
Air Blank	0.000	00:13
Control Test	0.081	00:13
Air Blank	0.000	00:14
Subject Sample # 1	0.134	00:15
Breath Volume	1.183	00:15
Air Blank	0.000	00:15
Air Blank	0.000	00:17
Subject Sample # 2	0.138	00:18
Breath Volume	1.882	00:18
Air Blank	0.000	00:18
Control Test	0.080	00:19
Air Blank	0.000	00:19
Diagnostics Check	OK	00:19

Data Download Date/Time: 10/24/2017 18:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LETSON,REBECCA,G
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: LETSON,REBECCA
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 09/26/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 03:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:34

SUBJECT'S LAST NAME: PARREA

DATE/TIME OF TEST: 10/14/2017 04:09:35

Diagnostics Check		OK	03:58	
Air Blank		0.000	03:58	
Control Test		0.080	03:58	
Air Blank		0.000	03:59	
Subject Sample	# 1	0.015	04:02	Volume Not Met
Breath Volume		0.203	04:02	
Air Blank		0.000	04:03	
Air Blank		0.000	04:04	
Subject Sample	# 2	0.011	04:08	Volume Not Met
Breath Volume		0.289	04:08	
Air Blank		0.000	04:08	
Control Test		0.078	04:08	
Air Blank		0.000	04:09	
Diagnostics Check		OK	04:09	

Data Download Date/Time: 10/24/2017 18:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAKER,MARK,J
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: BAKER,MARK
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 09/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 20:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:35

SUBJECT'S LAST NAME: CLEMENTS

DATE/TIME OF TEST: 10/14/2017 22:01:53

Diagnostics Check	OK	21:55
Air Blank	0.000	21:56
Control Test	0.080	21:56
Air Blank	0.000	21:56
Subject Sample # 1	0.199	21:57
Breath Volume	3.230	21:57
Air Blank	0.000	21:58
Air Blank	0.000	21:59
Subject Sample # 2	0.187	22:00
Breath Volume	3.128	22:00
Air Blank	0.000	22:00
Control Test	0.080	22:01
Air Blank	0.000	22:01
Diagnostics Check	OK	22:01

Data Download Date/Time: 10/24/2017 18:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERNANDEZ,ALAIN,
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: FERNANDEZ,ALAIN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 09/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 07:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:47

SUBJECT'S LAST NAME: AGUIRRE

DATE/TIME OF TEST: 10/15/2017 08:14:35

Diagnostics Check	OK	08:08
Air Blank	0.000	08:08
Control Test	0.080	08:09
Air Blank	0.000	08:09
Subject Sample # 1	0.048	08:10
Breath Volume	2.222	08:10
Air Blank	0.000	08:10
Air Blank	0.000	08:12
Subject Sample # 2	0.046	08:13
Breath Volume	1.289	08:13
Air Blank	0.000	08:13
Control Test	0.080	08:13
Air Blank	0.000	08:14
Diagnostics Check	OK	08:14

Data Download Date/Time: 10/24/2017 18:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARPENTER,DONALD,M
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: CARPENTER,DONALD
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 09/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 23:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: PATTERSON

DATE/TIME OF TEST: 10/16/2017 00:24:42

Diagnostics Check		OK	00:14
Air Blank		0.000	00:15
Control Test		0.080	00:15
Air Blank		0.000	00:15
Subject Sample	# 1	0.216	00:16
Breath Volume		1.734	00:16
Air Blank		0.000	00:17
Air Blank		0.000	00:18
Subject Sample	# 2	0.183	00:19
Breath Volume		1.101	00:19
Air Blank		0.000	00:20
Air Blank		0.000	00:22
Subject Sample	# 3	0.194	00:22
Breath Volume		1.296	00:22
Air Blank		0.000	00:23
Control Test		0.079	00:24
Air Blank		0.000	00:24
Diagnostics Check		OK	00:24

Data Download Date/Time: 10/24/2017 18:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AUSTIN,NICHOLAS,J
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: REED,R
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 09/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 02:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:37

SUBJECT'S LAST NAME: BANDA

DATE/TIME OF TEST: 10/16/2017 04:04:07

Diagnostics Check	OK	03:58
Air Blank	0.000	03:58
Control Test	0.080	03:58
Air Blank	0.000	03:59
Subject Sample # 1	0.156	03:59
Breath Volume	3.046	03:59
Air Blank	0.000	04:00
Air Blank	0.000	04:02
Subject Sample # 2	0.152	04:02
Breath Volume	3.367	04:02
Air Blank	0.000	04:03
Control Test	0.079	04:03
Air Blank	0.000	04:04
Diagnostics Check	OK	04:04

Data Download Date/Time: 10/24/2017 18:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCBRIDE,TARA,
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: MCBRIDE,TARA
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 09/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 09:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:20

SUBJECT'S LAST NAME: BENNETT

DATE/TIME OF TEST: 10/17/2017 10:59:40

Diagnostics Check	OK	10:52
Air Blank	0.000	10:53
Control Test	0.079	10:53
Air Blank	0.000	10:54
Subject Sample # 1	0.000	10:55
Breath Volume	2.191	10:55
Air Blank	0.000	10:55
Air Blank	0.000	10:57
Subject Sample # 2	0.000	10:58
Breath Volume	1.871	10:58
Air Blank	0.000	10:58
Control Test	0.078	10:59
Air Blank	0.000	10:59
Diagnostics Check	OK	10:59

Data Download Date/Time: 10/24/2017 18:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,GARY,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: MILLER,GARY
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 09/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 00:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:12

SUBJECT'S LAST NAME: RICHERSON

DATE/TIME OF TEST: 10/21/2017 01:43:44

Diagnostics Check		OK	01:37
Air Blank		0.000	01:37
Control Test		0.080	01:37
Air Blank		0.000	01:38
Subject Sample	# 1	0.216	01:39
Breath Volume		2.980	01:39
Air Blank		0.000	01:39
Air Blank		0.000	01:41
Subject Sample	# 2	0.225	01:42
Breath Volume		2.097	01:42
Air Blank		0.000	01:42
Control Test		0.079	01:43
Air Blank		0.000	01:43
Diagnostics Check		OK	01:43

Data Download Date/Time: 10/24/2017 18:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LACOGNATA,CHRISTINA,T
BREATH TEST OPERATOR AGENCY: Belleview PD
ARRESTING OFFICER: LACOGNATA,CHRISTINA
ARRESTING OFFICER AGENCY: Belleview PD
LAST AGENCY INSPECTION: 09/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 23:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: WARBLE

DATE/TIME OF TEST: 10/22/2017 00:59:58

Diagnostics Check	OK	00:55	
Air Blank	0.000	00:55	
Control Test	0.080	00:56	
Air Blank	0.000	00:56	
Subject Sample	# 1	00:59	Sequence Aborted
Breath Volume	0.121	00:59	
Air Blank	0.000	00:59	

Data Download Date/Time: 10/24/2017 18:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LACOGNATA,CHRISTINA,T
BREATH TEST OPERATOR AGENCY: Belleview PD
ARRESTING OFFICER: LACOGNATA,CHRISTINA
ARRESTING OFFICER AGENCY: Belleview PD
LAST AGENCY INSPECTION: 09/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 23:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: WARBLE

DATE/TIME OF TEST: 10/22/2017 01:05:33

Diagnostics Check	OK	01:01	
Air Blank	0.000	01:01	
Control Test	0.078	01:02	
Air Blank	0.000	01:02	
Subject Sample	# 1	01:04	Sequence Aborted
Breath Volume	0.218	01:04	
Air Blank	0.000	01:05	

Data Download Date/Time: 10/24/2017 18:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LACOGNATA,CHRISTINA,T
BREATH TEST OPERATOR AGENCY: Belleview PD
ARRESTING OFFICER: LACOGNATA,CHRISTINA
ARRESTING OFFICER AGENCY: Belleview PD
LAST AGENCY INSPECTION: 09/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 03:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:09

SUBJECT'S LAST NAME: COMPTON

DATE/TIME OF TEST: 10/23/2017 05:36:41

Diagnostics Check	OK	05:29
Air Blank	0.000	05:30
Control Test	0.080	05:30
Air Blank	0.000	05:31
Subject Sample # 1	0.169	05:31
Breath Volume	1.683	05:31
Air Blank	0.000	05:32
Air Blank	0.000	05:34
Subject Sample # 2	0.170	05:34
Breath Volume	2.609	05:34
Air Blank	0.000	05:35
Control Test	0.078	05:36
Air Blank	0.000	05:36
Diagnostics Check	OK	05:36

Data Download Date/Time: 10/24/2017 18:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DE JESUS,V
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:46

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 10/05/2017 00:09:20

Diagnostics Check	OK	00:06	
Air Blank	0.000	00:06	
Control Test	0.079	00:07	
Air Blank	0.000	00:07	
Subject Sample	# 1	00:07	Subject Test Refused
Breath Volume	0.000	00:07	
Air Blank	0.000	00:08	
Control Test	0.079	00:08	
Air Blank	0.000	00:09	
Diagnostics Check	OK	00:09	

Data Download Date/Time: 11/21/2017 14:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LYON,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 10/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:02

SUBJECT'S LAST NAME: ARANG

DATE/TIME OF TEST: 10/05/2017 05:28:17

Diagnostics Check	OK	05:22
Air Blank	0.000	05:22
Control Test	0.079	05:22
Air Blank	0.000	05:23
Subject Sample # 1	0.204	05:23
Breath Volume	2.484	05:23
Air Blank	0.000	05:24
Air Blank	0.000	05:26
Subject Sample # 2	0.191	05:26
Breath Volume	2.355	05:26
Air Blank	0.000	05:27
Control Test	0.082	05:27
Air Blank	0.000	05:28
Diagnostics Check	OK	05:28

Data Download Date/Time: 11/21/2017 14:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKIPPER,ROGER,G
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: KENNEDY,D 129956
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 08:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:11

SUBJECT'S LAST NAME: WALLACE

DATE/TIME OF TEST: 10/05/2017 09:34:59

Diagnostics Check	OK	09:31	
Air Blank	0.000	09:31	
Control Test	0.079	09:32	
Air Blank	0.000	09:32	
Subject Sample	# 1	09:34	Sequence Aborted
Breath Volume	0.000	09:34	
Air Blank	0.000	09:34	

Data Download Date/Time: 11/21/2017 14:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: TREMBLAY, J
ARRESTING OFFICER AGENCY: Tampa Airport PD
LAST AGENCY INSPECTION: 10/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 17:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:43

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 10/05/2017 21:10:11

Diagnostics Check		OK	21:04
Air Blank		0.000	21:04
Control Test		0.080	21:04
Air Blank		0.000	21:05
Subject Sample	# 1	0.204	21:05
Breath Volume		2.550	21:05
Air Blank		0.000	21:06
Air Blank		0.000	21:08
Subject Sample	# 2	0.201	21:08
Breath Volume		1.695	21:08
Air Blank		0.000	21:09
Control Test		0.079	21:09
Air Blank		0.000	21:10
Diagnostics Check		OK	21:10

Data Download Date/Time: 11/21/2017 14:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARLAUG, G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 10/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:14

SUBJECT'S LAST NAME: GOMEZ

DATE/TIME OF TEST: 10/06/2017 03:41:54

Diagnostics Check	OK	03:35
Air Blank	0.000	03:35
Control Test	0.078	03:36
Air Blank	0.000	03:36
Subject Sample # 1	0.126	03:37
Breath Volume	1.863	03:37
Air Blank	0.000	03:38
Air Blank	0.000	03:40
Subject Sample # 2	0.114	03:40
Breath Volume	1.699	03:40
Air Blank	0.000	03:41
Control Test	0.078	03:41
Air Blank	0.000	03:41
Diagnostics Check	OK	03:41

Data Download Date/Time: 11/21/2017 14:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRIMBLE, TERESA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VAN TREESE, S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 10/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:48

SUBJECT'S LAST NAME: JAMES

DATE/TIME OF TEST: 10/06/2017 23:16:05

Diagnostics Check	OK	23:09
Air Blank	0.000	23:09
Control Test	0.079	23:10
Air Blank	0.000	23:10
Subject Sample # 1	0.283	23:11
Breath Volume	1.957	23:11
Air Blank	0.000	23:11
Air Blank	0.000	23:13
Subject Sample # 2	0.267	23:14
Breath Volume	1.179	23:14
Air Blank	0.000	23:15
Control Test	0.079	23:15
Air Blank	0.000	23:16
Diagnostics Check	OK	23:16

Data Download Date/Time: 11/21/2017 14:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SHEA, D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:13

SUBJECT'S LAST NAME: ACEVEDO REYNA

DATE/TIME OF TEST: 10/07/2017 01:41:43

Diagnostics Check	OK	01:34
Air Blank	0.000	01:34
Control Test	0.079	01:35
Air Blank	0.000	01:35
Subject Sample # 1	0.156	01:37
Breath Volume	1.660	01:37
Air Blank	0.000	01:37
Air Blank	0.000	01:39
Subject Sample # 2	0.159	01:40
Breath Volume	1.429	01:40
Air Blank	0.000	01:40
Control Test	0.078	01:41
Air Blank	0.000	01:41
Diagnostics Check	OK	01:41

Data Download Date/Time: 11/21/2017 14:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PAULINO,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:46

SUBJECT'S LAST NAME: CORDERO CRUZ

DATE/TIME OF TEST: 10/07/2017 04:16:19

Diagnostics Check	OK	04:06
Air Blank	0.000	04:06
Control Test	0.079	04:07
Air Blank	0.000	04:07
Subject Sample # 1	0.224	04:08
Breath Volume	1.343	04:08
Air Blank	0.000	04:08
Air Blank	0.000	04:10
Subject Sample # 2	0.258	04:11
Breath Volume	2.535	04:11
Air Blank	0.000	04:12
Air Blank	0.000	04:14
Subject Sample # 3	0.264	04:14
Breath Volume	2.917	04:14
Air Blank	0.000	04:15
Control Test	0.078	04:15
Air Blank	0.000	04:16
Diagnostics Check	OK	04:16

Data Download Date/Time: 11/21/2017 14:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SIMPKINS, D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:34

SUBJECT'S LAST NAME: POLANCO

DATE/TIME OF TEST: 10/07/2017 05:03:30

Diagnostics Check	OK	04:56
Air Blank	0.000	04:56
Control Test	0.079	04:57
Air Blank	0.000	04:57
Subject Sample # 1	0.268	04:58
Breath Volume	1.632	04:58
Air Blank	0.000	04:59
Air Blank	0.000	05:01
Subject Sample # 2	0.249	05:01
Breath Volume	1.660	05:01
Air Blank	0.000	05:02
Control Test	0.078	05:02
Air Blank	0.000	05:03
Diagnostics Check	OK	05:03

Data Download Date/Time: 11/21/2017 14:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRIMBLE, TERESA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRABAND, M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 10/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:14

SUBJECT'S LAST NAME: TWIETMEYER

DATE/TIME OF TEST: 10/07/2017 05:41:18

Diagnostics Check	OK	05:34
Air Blank	0.000	05:35
Control Test	0.079	05:35
Air Blank	0.000	05:36
Subject Sample # 1	0.188	05:36
Breath Volume	2.664	05:36
Air Blank	0.000	05:37
Air Blank	0.000	05:39
Subject Sample # 2	0.185	05:39
Breath Volume	2.105	05:39
Air Blank	0.000	05:40
Control Test	0.079	05:40
Air Blank	0.000	05:41
Diagnostics Check	OK	05:41

Data Download Date/Time: 11/21/2017 14:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNYDER,MELONIE,E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VELAZQUEZ,B
ARRESTING OFFICER AGENCY: U of So FL Campus PD
LAST AGENCY INSPECTION: 10/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 07:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:32

SUBJECT'S LAST NAME: ZHU

DATE/TIME OF TEST: 10/07/2017 08:58:52

Diagnostics Check	OK	08:52
Air Blank	0.000	08:53
Control Test	0.079	08:53
Air Blank	0.000	08:54
Subject Sample # 1	0.000	08:54
Breath Volume	1.742	08:54
Air Blank	0.000	08:55
Air Blank	0.000	08:57
Subject Sample # 2	0.000	08:57
Breath Volume	2.000	08:57
Air Blank	0.000	08:57
Control Test	0.078	08:58
Air Blank	0.000	08:58
Diagnostics Check	OK	08:58

Data Download Date/Time: 11/21/2017 14:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 10/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:54

SUBJECT'S LAST NAME: SUAREZ ESTEVEZ

DATE/TIME OF TEST: 10/07/2017 22:23:01

Diagnostics Check	OK	22:14
Air Blank	0.000	22:15
Control Test	0.079	22:15
Air Blank	0.000	22:15
Subject Sample # 1	0.117	22:18
Breath Volume	1.855	22:18
Air Blank	0.000	22:18
Air Blank	0.000	22:20
Subject Sample # 2	0.119	22:21
Breath Volume	1.820	22:21
Air Blank	0.000	22:22
Control Test	0.077	22:22
Air Blank	0.000	22:22
Diagnostics Check	OK	22:23

Data Download Date/Time: 11/21/2017 14:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: CRAGG,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 10/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:53

SUBJECT'S LAST NAME: LOPEZ-SEMPO

DATE/TIME OF TEST: 10/08/2017 04:20:13

Diagnostics Check	OK	04:13
Air Blank	0.000	04:14
Control Test	0.079	04:14
Air Blank	0.000	04:15
Subject Sample # 1	0.112	04:15
Breath Volume	1.183	04:15
Air Blank	0.000	04:16
Air Blank	0.000	04:18
Subject Sample # 2	0.121	04:18
Breath Volume	1.304	04:18
Air Blank	0.000	04:19
Control Test	0.079	04:19
Air Blank	0.000	04:20
Diagnostics Check	OK	04:20

Data Download Date/Time: 11/21/2017 14:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,REBECCA,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RAMOS,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 06:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:42

SUBJECT'S LAST NAME: WADSWORTH

DATE/TIME OF TEST: 10/08/2017 08:10:12

Diagnostics Check		OK	08:03
Air Blank		0.000	08:03
Control Test		0.079	08:03
Air Blank		0.000	08:04
Subject Sample	# 1	0.214	08:05
Breath Volume		1.171	08:05
Air Blank		0.000	08:06
Air Blank		0.000	08:07
Subject Sample	# 2	0.218	08:08
Breath Volume		1.703	08:08
Air Blank		0.000	08:09
Control Test		0.078	08:09
Air Blank		0.000	08:10
Diagnostics Check		OK	08:10

Data Download Date/Time: 11/21/2017 14:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LARA,A
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 10/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:39

SUBJECT'S LAST NAME: AVILES-RUIZ

DATE/TIME OF TEST: 10/08/2017 23:08:29

Diagnostics Check		OK	22:59
Air Blank		0.000	22:59
Control Test		0.079	23:00
Air Blank		0.000	23:00
Subject Sample	# 1	0.221	23:03
Breath Volume		1.261	23:03
Air Blank		0.000	23:03
Air Blank		0.000	23:05
Subject Sample	# 2	0.230	23:06
Breath Volume		1.265	23:06
Air Blank		0.000	23:07
Control Test		0.078	23:07
Air Blank		0.000	23:08
Diagnostics Check		OK	23:08

Data Download Date/Time: 11/21/2017 14:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 10/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:21

SUBJECT'S LAST NAME: SHEA

DATE/TIME OF TEST: 10/09/2017 22:47:53

Diagnostics Check	OK	22:41
Air Blank	0.000	22:42
Control Test	0.079	22:42
Air Blank	0.000	22:42
Subject Sample # 1	0.057	22:43
Breath Volume	1.242	22:43
Air Blank	0.000	22:43
Air Blank	0.000	22:45
Subject Sample # 2	0.054	22:46
Breath Volume	1.292	22:46
Air Blank	0.000	22:46
Control Test	0.079	22:47
Air Blank	0.000	22:47
Diagnostics Check	OK	22:47

Data Download Date/Time: 11/21/2017 14:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MACIAS, A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: FISH

DATE/TIME OF TEST: 10/11/2017 00:32:24

Diagnostics Check	OK	00:25
Air Blank	0.000	00:26
Control Test	0.079	00:26
Air Blank	0.000	00:27
Subject Sample # 1	0.172	00:27
Breath Volume	1.792	00:27
Air Blank	0.000	00:28
Air Blank	0.000	00:30
Subject Sample # 2	0.167	00:30
Breath Volume	1.660	00:30
Air Blank	0.000	00:31
Control Test	0.078	00:31
Air Blank	0.000	00:32
Diagnostics Check	OK	00:32

Data Download Date/Time: 11/21/2017 14:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRIMBLE, TERESA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRENNAN, D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:04

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 10/11/2017 04:31:38

Diagnostics Check	OK	04:24
Air Blank	0.000	04:25
Control Test	0.079	04:25
Air Blank	0.000	04:26
Subject Sample # 1	0.202	04:27
Breath Volume	1.765	04:27
Air Blank	0.000	04:27
Air Blank	0.000	04:29
Subject Sample # 2	0.203	04:30
Breath Volume	1.464	04:30
Air Blank	0.000	04:30
Control Test	0.078	04:31
Air Blank	0.000	04:31
Diagnostics Check	OK	04:31

Data Download Date/Time: 11/21/2017 14:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SIMPKINS, D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:07

SUBJECT'S LAST NAME: DAVIS II

DATE/TIME OF TEST: 10/11/2017 23:31:36

Diagnostics Check	OK	23:28	
Air Blank	0.000	23:29	
Control Test	0.079	23:29	
Air Blank	0.000	23:30	
Subject Sample	# 1	23:30	Subject Test Refused
Breath Volume	0.000	23:30	
Air Blank	0.000	23:30	
Control Test	0.079	23:31	
Air Blank	0.000	23:31	
Diagnostics Check	OK	23:31	

Data Download Date/Time: 11/21/2017 14:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: NIEUWENDAM,R
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 10/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:02

SUBJECT'S LAST NAME: CROSBY

DATE/TIME OF TEST: 10/12/2017 01:29:11

Diagnostics Check	OK	01:22
Air Blank	0.000	01:23
Control Test	0.079	01:23
Air Blank	0.000	01:24
Subject Sample # 1	0.000	01:24
Breath Volume	2.410	01:24
Air Blank	0.000	01:25
Air Blank	0.000	01:27
Subject Sample # 2	0.000	01:27
Breath Volume	2.324	01:27
Air Blank	0.000	01:28
Control Test	0.079	01:28
Air Blank	0.000	01:29
Diagnostics Check	OK	01:29

Data Download Date/Time: 11/21/2017 14:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SHEA,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 19:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:02

SUBJECT'S LAST NAME: COCKERHAM

DATE/TIME OF TEST: 10/12/2017 20:29:05

Diagnostics Check	OK	20:22
Air Blank	0.000	20:23
Control Test	0.079	20:23
Air Blank	0.000	20:23
Subject Sample # 1	0.163	20:24
Breath Volume	2.605	20:24
Air Blank	0.000	20:25
Air Blank	0.000	20:27
Subject Sample # 2	0.168	20:27
Breath Volume	3.007	20:27
Air Blank	0.000	20:28
Control Test	0.078	20:28
Air Blank	0.000	20:29
Diagnostics Check	OK	20:29

Data Download Date/Time: 11/21/2017 14:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ORNDORFF,W
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:03

SUBJECT'S LAST NAME: TORRES

DATE/TIME OF TEST: 10/13/2017 04:36:32

Diagnostics Check		OK	04:24	
Air Blank		0.000	04:25	
Control Test		0.080	04:25	
Air Blank		0.000	04:26	
Subject Sample	# 1	0.122	04:29	Volume Not Met
Breath Volume		0.214	04:29	
Air Blank		0.000	04:30	
Air Blank		0.000	04:31	
Subject Sample	# 2	0.132	04:34	Volume Not Met
Breath Volume		0.097	04:34	
Air Blank		0.000	04:35	
Control Test		0.078	04:35	
Air Blank		0.000	04:36	
Diagnostics Check		OK	04:36	

Data Download Date/Time: 11/21/2017 14:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK,CASEY,S
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DAWSY,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:58

SUBJECT'S LAST NAME: SAVARY

DATE/TIME OF TEST: 10/13/2017 21:30:32

Diagnostics Check		OK	21:18	
Air Blank		0.000	21:19	
Control Test		0.079	21:19	
Air Blank		0.000	21:20	
Subject Sample	# 1	0.158	21:23	Volume Not Met
Breath Volume		0.046	21:23	
Air Blank		0.000	21:23	
Air Blank		0.000	21:25	
Subject Sample	# 2	0.176	21:28	Volume Not Met
Breath Volume		0.050	21:28	
Air Blank		0.000	21:29	
Control Test		0.078	21:29	
Air Blank		0.000	21:30	
Diagnostics Check		OK	21:30	

Data Download Date/Time: 11/21/2017 14:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARLAUG,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 10/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:11

SUBJECT'S LAST NAME: RONSO

DATE/TIME OF TEST: 10/14/2017 04:35:00

Diagnostics Check	OK	04:31	
Air Blank	0.000	04:32	
Control Test	0.079	04:32	
Air Blank	0.000	04:33	
Subject Sample	# 1	04:33	Subject Test Refused
Breath Volume	0.000	04:33	
Air Blank	0.000	04:34	
Control Test	0.078	04:34	
Air Blank	0.000	04:34	
Diagnostics Check	OK	04:35	

Data Download Date/Time: 11/21/2017 14:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRIMBLE, TERESA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRENNAN, D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 19:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:44

SUBJECT'S LAST NAME: LOPEZ LOPEZ

DATE/TIME OF TEST: 10/14/2017 20:12:05

Diagnostics Check	OK	20:04
Air Blank	0.000	20:05
Control Test	0.078	20:05
Air Blank	0.000	20:05
Subject Sample # 1	0.203	20:06
Breath Volume	1.281	20:06
Air Blank	0.000	20:07
Air Blank	0.000	20:09
Subject Sample # 2	0.208	20:10
Breath Volume	1.515	20:10
Air Blank	0.000	20:11
Control Test	0.079	20:11
Air Blank	0.000	20:12
Diagnostics Check	OK	20:12

Data Download Date/Time: 11/21/2017 14:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRENNAN, D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:18

SUBJECT'S LAST NAME: INVERARY

DATE/TIME OF TEST: 10/15/2017 02:51:26

Diagnostics Check		OK	02:38	
Air Blank		0.000	02:39	
Control Test		0.079	02:39	
Air Blank		0.000	02:40	
Subject Sample	# 1	0.048	02:43	Volume Not Met
Breath Volume		0.542	02:43	
Air Blank		0.000	02:43	
Air Blank		0.000	02:45	
Subject Sample	# 2	0.059	02:46	
Breath Volume		1.816	02:46	
Air Blank		0.000	02:47	
Air Blank		0.000	02:49	
Subject Sample	# 3	0.053	02:49	
Breath Volume		2.218	02:49	
Air Blank		0.000	02:50	
Control Test		0.078	02:50	
Air Blank		0.000	02:51	
Diagnostics Check		OK	02:51	

Data Download Date/Time: 11/21/2017 14:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRIMBLE, TERESA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LYON, M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 10/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:56

SUBJECT'S LAST NAME: ALICEA PADILLA

DATE/TIME OF TEST: 10/15/2017 04:27:44

Diagnostics Check	OK	04:17
Air Blank	0.000	04:17
Control Test	0.079	04:17
Air Blank	0.000	04:18
Subject Sample # 1	0.166	04:21
Breath Volume	1.394	04:21
Air Blank	0.000	04:21
Air Blank	0.000	04:23
Subject Sample # 2	0.170	04:26
Breath Volume	2.839	04:26
Air Blank	0.000	04:26
Control Test	0.078	04:27
Air Blank	0.000	04:27
Diagnostics Check	OK	04:27

Data Download Date/Time: 11/21/2017 14:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FRANKLIN, R
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:48

SUBJECT'S LAST NAME: SHRAUDER

DATE/TIME OF TEST: 10/15/2017 05:15:38

Diagnostics Check	OK	05:08
Air Blank	0.000	05:09
Control Test	0.079	05:09
Air Blank	0.000	05:10
Subject Sample # 1	0.196	05:11
Breath Volume	1.378	05:11
Air Blank	0.000	05:11
Air Blank	0.000	05:13
Subject Sample # 2	0.200	05:14
Breath Volume	1.207	05:14
Air Blank	0.000	05:14
Control Test	0.078	05:15
Air Blank	0.000	05:15
Diagnostics Check	OK	05:15

Data Download Date/Time: 11/21/2017 14:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRIMBLE, TERESA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MONTOUTE, S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 05:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:40

SUBJECT'S LAST NAME: TOMLINSON

DATE/TIME OF TEST: 10/15/2017 06:07:00

Diagnostics Check		OK	06:00
Air Blank		0.000	06:01
Control Test		0.078	06:01
Air Blank		0.000	06:01
Subject Sample	# 1	0.153	06:02
Breath Volume		2.480	06:02
Air Blank		0.000	06:02
Air Blank		0.000	06:04
Subject Sample	# 2	0.153	06:05
Breath Volume		2.375	06:05
Air Blank		0.000	06:06
Control Test		0.078	06:06
Air Blank		0.000	06:06
Diagnostics Check		OK	06:07

Data Download Date/Time: 11/21/2017 14:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMMONS,ANGELA,V
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DEJESUS,VICTOR
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 06:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:54

SUBJECT'S LAST NAME: RUIZ-NUNEZ

DATE/TIME OF TEST: 10/15/2017 07:21:49

Diagnostics Check	OK	07:15
Air Blank	0.000	07:15
Control Test	0.079	07:15
Air Blank	0.000	07:16
Subject Sample # 1	0.127	07:17
Breath Volume	2.808	07:17
Air Blank	0.000	07:17
Air Blank	0.000	07:19
Subject Sample # 2	0.112	07:20
Breath Volume	2.640	07:20
Air Blank	0.000	07:20
Control Test	0.078	07:21
Air Blank	0.000	07:21
Diagnostics Check	OK	07:21

Data Download Date/Time: 11/21/2017 14:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SINANI,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:43

SUBJECT'S LAST NAME: NUNEZ MEDINA

DATE/TIME OF TEST: 10/15/2017 21:16:45

Diagnostics Check		OK	21:04	
Air Blank		0.000	21:04	
Control Test		0.080	21:05	
Air Blank		0.000	21:05	
Subject Sample	# 1	0.152	21:08	Volume Not Met
Breath Volume		0.636	21:08	
Air Blank		0.000	21:09	
Air Blank		0.000	21:11	
Subject Sample	# 2	0.181	21:11	
Breath Volume		1.398	21:11	
Air Blank		0.000	21:12	
Air Blank		0.000	21:14	
Subject Sample	# 3	0.176	21:15	
Breath Volume		1.214	21:15	
Air Blank		0.000	21:15	
Control Test		0.077	21:16	
Air Blank		0.000	21:16	
Diagnostics Check		OK	21:16	

Data Download Date/Time: 11/21/2017 14:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRIMBLE, TERESA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FOWLER, J
ARRESTING OFFICER AGENCY: Plant City PD
LAST AGENCY INSPECTION: 10/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:39

SUBJECT'S LAST NAME: SOTO

DATE/TIME OF TEST: 10/15/2017 23:09:59

Diagnostics Check		OK	22:59
Air Blank		0.000	23:00
Control Test		0.079	23:00
Air Blank		0.000	23:01
Subject Sample	# 1	0.273	23:02
Breath Volume		1.242	23:02
Air Blank		0.000	23:02
Air Blank		0.000	23:04
Subject Sample	# 2	0.295	23:05
Breath Volume		2.109	23:05
Air Blank		0.000	23:05
Air Blank		0.000	23:07
Subject Sample	# 3	0.301	23:08
Breath Volume		2.269	23:08
Air Blank		0.000	23:09
Control Test		0.078	23:09
Air Blank		0.000	23:09
Diagnostics Check		OK	23:09

Data Download Date/Time: 11/21/2017 14:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DUNCAN,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:17

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 10/16/2017 22:44:08

Diagnostics Check	OK	22:37
Air Blank	0.000	22:37
Control Test	0.079	22:38
Air Blank	0.000	22:38
Subject Sample # 1	0.193	22:39
Breath Volume	1.687	22:39
Air Blank	0.000	22:40
Air Blank	0.000	22:42
Subject Sample # 2	0.196	22:42
Breath Volume	1.742	22:42
Air Blank	0.000	22:43
Control Test	0.078	22:43
Air Blank	0.000	22:44
Diagnostics Check	OK	22:44

Data Download Date/Time: 11/21/2017 14:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS, S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 10/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:21

SUBJECT'S LAST NAME: ROQUE PEREZ

DATE/TIME OF TEST: 10/17/2017 02:50:49

Diagnostics Check	OK	02:42	
Air Blank	0.000	02:42	
Control Test	0.078	02:43	
Air Blank	0.000	02:43	
Subject Sample	# 1	02:46	No Sample Provided
Breath Volume	0.000	02:46	
Air Blank	0.000	02:47	
Air Blank	0.000	02:49	
Subject Sample	# 2	02:49	Subject Test Refused
Breath Volume	0.000	02:49	
Air Blank	0.000	02:49	
Control Test	0.078	02:50	
Air Blank	0.000	02:50	
Diagnostics Check	OK	02:50	

Data Download Date/Time: 11/21/2017 14:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNYDER,MELONIE,E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARLOCK,R
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 10:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:56

SUBJECT'S LAST NAME: RUFF

DATE/TIME OF TEST: 10/17/2017 11:22:51

Diagnostics Check	OK	11:16
Air Blank	0.000	11:16
Control Test	0.079	11:17
Air Blank	0.000	11:17
Subject Sample # 1	0.000	11:18
Breath Volume	2.906	11:18
Air Blank	0.000	11:19
Air Blank	0.000	11:21
Subject Sample # 2	0.000	11:21
Breath Volume	2.750	11:21
Air Blank	0.000	11:21
Control Test	0.079	11:22
Air Blank	0.000	11:22
Diagnostics Check	OK	11:22

Data Download Date/Time: 11/21/2017 14:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PORTMAN,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 10/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: MCKENNA

DATE/TIME OF TEST: 10/18/2017 02:57:25

Diagnostics Check	OK	02:50
Air Blank	0.000	02:51
Control Test	0.079	02:51
Air Blank	0.000	02:52
Subject Sample # 1	0.177	02:52
Breath Volume	1.378	02:52
Air Blank	0.000	02:53
Air Blank	0.000	02:55
Subject Sample # 2	0.182	02:55
Breath Volume	1.382	02:55
Air Blank	0.000	02:56
Control Test	0.079	02:56
Air Blank	0.000	02:57
Diagnostics Check	OK	02:57

Data Download Date/Time: 11/21/2017 14:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BLANCHARD,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 10/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 05:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:30

SUBJECT'S LAST NAME: MCCARTHY

DATE/TIME OF TEST: 10/18/2017 05:57:22

Diagnostics Check	OK	05:50
Air Blank	0.000	05:50
Control Test	0.080	05:50
Air Blank	0.000	05:51
Subject Sample # 1	0.164	05:52
Breath Volume	1.328	05:52
Air Blank	0.000	05:53
Air Blank	0.000	05:54
Subject Sample # 2	0.157	05:55
Breath Volume	1.398	05:55
Air Blank	0.000	05:56
Control Test	0.079	05:56
Air Blank	0.000	05:57
Diagnostics Check	OK	05:57

Data Download Date/Time: 11/21/2017 14:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SINANI,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: HESS

DATE/TIME OF TEST: 10/19/2017 00:29:52

Diagnostics Check		OK	00:15	
Air Blank		0.000	00:15	
Control Test		0.079	00:16	
Air Blank		0.000	00:16	
Subject Sample	# 1	0.161	00:19	Volume Not Met
Breath Volume		0.781	00:19	
Air Blank		0.000	00:20	
Air Blank		0.000	00:22	
Subject Sample	# 2	0.187	00:24	
Breath Volume		1.187	00:24	
Air Blank		0.000	00:25	
Air Blank		0.000	00:27	
Subject Sample	# 3	0.191	00:28	
Breath Volume		1.152	00:28	
Air Blank		0.000	00:28	
Control Test		0.078	00:29	
Air Blank		0.000	00:29	
Diagnostics Check		OK	00:29	

Data Download Date/Time: 11/21/2017 14:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS, S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 10/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 10/20/2017 01:12:49

Diagnostics Check	OK	01:05
Air Blank	0.000	01:06
Control Test	0.079	01:06
Air Blank	0.000	01:07
Subject Sample # 1	0.175	01:08
Breath Volume	2.769	01:08
Air Blank	0.000	01:08
Air Blank	0.000	01:10
Subject Sample # 2	0.167	01:11
Breath Volume	1.531	01:11
Air Blank	0.000	01:11
Control Test	0.078	01:12
Air Blank	0.000	01:12
Diagnostics Check	OK	01:12

Data Download Date/Time: 11/21/2017 14:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMMONS,ANGELA,V
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARCIA,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 17:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:13

SUBJECT'S LAST NAME: COCKERHAM

DATE/TIME OF TEST: 10/20/2017 18:39:59

Diagnostics Check	OK	18:33
Air Blank	0.000	18:33
Control Test	0.079	18:34
Air Blank	0.000	18:34
Subject Sample # 1	0.133	18:35
Breath Volume	2.027	18:35
Air Blank	0.000	18:35
Air Blank	0.000	18:37
Subject Sample # 2	0.139	18:38
Breath Volume	2.246	18:38
Air Blank	0.000	18:39
Control Test	0.079	18:39
Air Blank	0.000	18:39
Diagnostics Check	OK	18:39

Data Download Date/Time: 11/21/2017 14:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRABAND, M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 10/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:14

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 10/21/2017 04:46:42

Diagnostics Check	OK	04:37
Air Blank	0.000	04:38
Control Test	0.079	04:38
Air Blank	0.000	04:39
Subject Sample # 1	0.197	04:39
Breath Volume	1.738	04:39
Air Blank	0.000	04:40
Air Blank	0.000	04:42
Subject Sample # 2	0.211	04:45
Breath Volume	1.390	04:45
Air Blank	0.000	04:45
Control Test	0.080	04:46
Air Blank	0.000	04:46
Diagnostics Check	OK	04:46

Data Download Date/Time: 11/21/2017 14:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DUNCAN,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:54

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 10/21/2017 05:21:57

Diagnostics Check	OK	05:15
Air Blank	0.000	05:15
Control Test	0.079	05:16
Air Blank	0.000	05:16
Subject Sample # 1	0.219	05:17
Breath Volume	1.503	05:17
Air Blank	0.000	05:17
Air Blank	0.000	05:19
Subject Sample # 2	0.206	05:20
Breath Volume	1.316	05:20
Air Blank	0.000	05:21
Control Test	0.078	05:21
Air Blank	0.000	05:21
Diagnostics Check	OK	05:21

Data Download Date/Time: 11/21/2017 14:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:43

SUBJECT'S LAST NAME: GONZALEZ-ORTIZ

DATE/TIME OF TEST: 10/22/2017 00:10:07

Diagnostics Check	OK	00:03
Air Blank	0.000	00:03
Control Test	0.080	00:04
Air Blank	0.000	00:04
Subject Sample # 1	0.151	00:05
Breath Volume	1.812	00:05
Air Blank	0.000	00:06
Air Blank	0.000	00:08
Subject Sample # 2	0.139	00:08
Breath Volume	2.007	00:08
Air Blank	0.000	00:09
Control Test	0.079	00:09
Air Blank	0.000	00:10
Diagnostics Check	OK	00:10

Data Download Date/Time: 11/21/2017 14:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRABAND,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 10/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:04

SUBJECT'S LAST NAME: FABIAN

DATE/TIME OF TEST: 10/22/2017 03:30:58

Diagnostics Check	OK	03:24
Air Blank	0.000	03:25
Control Test	0.080	03:25
Air Blank	0.000	03:25
Subject Sample # 1	0.194	03:26
Breath Volume	2.011	03:26
Air Blank	0.000	03:27
Air Blank	0.000	03:28
Subject Sample # 2	0.185	03:29
Breath Volume	1.125	03:29
Air Blank	0.000	03:30
Control Test	0.082	03:30
Air Blank	0.000	03:30
Diagnostics Check	OK	03:30

Data Download Date/Time: 11/21/2017 14:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: HERNANDEZ-VARGAS

DATE/TIME OF TEST: 10/22/2017 04:31:43

Diagnostics Check	OK	04:25
Air Blank	0.000	04:26
Control Test	0.080	04:26
Air Blank	0.000	04:26
Subject Sample # 1	0.194	04:27
Breath Volume	3.363	04:27
Air Blank	0.000	04:27
Air Blank	0.000	04:29
Subject Sample # 2	0.204	04:30
Breath Volume	3.242	04:30
Air Blank	0.000	04:30
Control Test	0.078	04:31
Air Blank	0.000	04:31
Diagnostics Check	OK	04:31

Data Download Date/Time: 11/21/2017 14:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SIMPKINS,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:36

SUBJECT'S LAST NAME: LARUE

DATE/TIME OF TEST: 10/23/2017 00:03:57

Diagnostics Check	OK	23:56
Air Blank	0.000	23:57
Control Test	0.080	23:57
Air Blank	0.000	23:58
Subject Sample # 2	0.172	23:58
Breath Volume	1.488	23:58
Air Blank	0.000	23:59
Air Blank	0.000	00:01
Subject Sample # 1	0.187	00:02
Breath Volume	2.363	00:02
Air Blank	0.000	00:03
Control Test	0.079	00:03
Air Blank	0.000	00:03
Diagnostics Check	OK	00:03

Data Download Date/Time: 11/21/2017 14:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SHEINBERG,T
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 10/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: CASE JR

DATE/TIME OF TEST: 10/23/2017 00:24:47

Diagnostics Check	OK	00:18
Air Blank	0.000	00:19
Control Test	0.080	00:19
Air Blank	0.000	00:19
Subject Sample # 1	0.208	00:20
Breath Volume	1.828	00:20
Air Blank	0.000	00:21
Air Blank	0.000	00:22
Subject Sample # 2	0.203	00:23
Breath Volume	1.515	00:23
Air Blank	0.000	00:23
Control Test	0.079	00:24
Air Blank	0.000	00:24
Diagnostics Check	OK	00:24

Data Download Date/Time: 11/21/2017 14:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FRANKLIN, R
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:22

SUBJECT'S LAST NAME: GALLAGHER

DATE/TIME OF TEST: 10/24/2017 21:48:31

Diagnostics Check	OK	21:42
Air Blank	0.000	21:42
Control Test	0.079	21:43
Air Blank	0.000	21:43
Subject Sample # 1	0.282	21:44
Breath Volume	5.222	21:44
Air Blank	0.000	21:44
Air Blank	0.000	21:46
Subject Sample # 2	0.284	21:47
Breath Volume	4.308	21:47
Air Blank	0.000	21:47
Control Test	0.078	21:47
Air Blank	0.000	21:48
Diagnostics Check	OK	21:48

Data Download Date/Time: 11/21/2017 14:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: NONE,NA
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: ADMIN
ARREST TIME: 22:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:52

SUBJECT'S LAST NAME: GELIS

DATE/TIME OF TEST: 10/25/2017 23:16:16

Diagnostics Check	OK	23:14	
Air Blank	0.000	23:14	
Control Test	0.080	23:15	
Air Blank	0.000	23:15	
		23:15	Improper Sample
Air Blank	0.000	23:16	

Data Download Date/Time: 11/21/2017 14:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: NONE, NA
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: ADMIN
ARREST TIME: 22:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: GELIS

DATE/TIME OF TEST: 10/25/2017 23:23:15

Diagnostics Check	OK	23:17
Air Blank	0.000	23:17
Control Test	0.080	23:18
Air Blank	0.000	23:18
Subject Sample # 1	0.057	23:18
Breath Volume	2.484	23:18
Air Blank	0.000	23:19
Air Blank	0.000	23:21
Subject Sample # 2	0.057	23:21
Breath Volume	2.253	23:21
Air Blank	0.000	23:22
Control Test	0.079	23:22
Air Blank	0.000	23:23
Diagnostics Check	OK	23:23

Data Download Date/Time: 11/21/2017 14:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: NIEUWENDAM,R
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 10/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:13

SUBJECT'S LAST NAME: MCDONALD

DATE/TIME OF TEST: 10/26/2017 01:41:20

Diagnostics Check	OK	01:33
Air Blank	0.000	01:34
Control Test	0.080	01:34
Air Blank	0.000	01:35
Subject Sample # 1	0.143	01:36
Breath Volume	1.300	01:36
Air Blank	0.000	01:37
Air Blank	0.000	01:39
Subject Sample # 2	0.156	01:39
Breath Volume	1.441	01:39
Air Blank	0.000	01:40
Control Test	0.079	01:40
Air Blank	0.000	01:41
Diagnostics Check	OK	01:41

Data Download Date/Time: 11/21/2017 14:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 10/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:37

SUBJECT'S LAST NAME: REYNOLDS

DATE/TIME OF TEST: 10/27/2017 02:04:52

Diagnostics Check		OK	01:57
Air Blank		0.000	01:58
Control Test		0.080	01:58
Air Blank		0.000	01:59
Subject Sample	# 1	0.109	02:00
Breath Volume		1.644	02:00
Air Blank		0.000	02:00
Air Blank		0.000	02:02
Subject Sample	# 2	0.119	02:03
Breath Volume		1.703	02:03
Air Blank		0.000	02:03
Control Test		0.080	02:04
Air Blank		0.000	02:04
Diagnostics Check		OK	02:04

Data Download Date/Time: 11/21/2017 14:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:34

SUBJECT'S LAST NAME: BARTLETT

DATE/TIME OF TEST: 10/27/2017 05:01:00

Diagnostics Check	OK	04:54
Air Blank	0.000	04:55
Control Test	0.080	04:55
Air Blank	0.000	04:56
Subject Sample # 1	0.155	04:56
Breath Volume	3.281	04:56
Air Blank	0.000	04:57
Air Blank	0.000	04:58
Subject Sample # 2	0.161	04:59
Breath Volume	4.027	04:59
Air Blank	0.000	05:00
Control Test	0.080	05:00
Air Blank	0.000	05:00
Diagnostics Check	OK	05:01

Data Download Date/Time: 11/21/2017 14:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRENNAN,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 19:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:28

SUBJECT'S LAST NAME: HAYES

DATE/TIME OF TEST: 10/27/2017 19:55:00

Diagnostics Check	OK	19:48
Air Blank	0.000	19:48
Control Test	0.080	19:49
Air Blank	0.000	19:49
Subject Sample # 1	0.147	19:50
Breath Volume	2.171	19:50
Air Blank	0.000	19:50
Air Blank	0.000	19:52
Subject Sample # 2	0.155	19:53
Breath Volume	3.300	19:53
Air Blank	0.000	19:54
Control Test	0.078	19:54
Air Blank	0.000	19:54
Diagnostics Check	OK	19:55

Data Download Date/Time: 11/21/2017 14:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: NIEUWENDAM,R
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 10/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:22

SUBJECT'S LAST NAME: JACKSON III

DATE/TIME OF TEST: 10/28/2017 01:52:20

Diagnostics Check	OK	01:42
Air Blank	0.000	01:42
Control Test	0.079	01:43
Air Blank	0.000	01:43
Subject Sample # 1	0.123	01:45
Breath Volume	1.281	01:45
Air Blank	0.000	01:46
Air Blank	0.000	01:48
Subject Sample # 2	0.113	01:50
Breath Volume	1.242	01:50
Air Blank	0.000	01:51
Control Test	0.078	01:51
Air Blank	0.000	01:52
Diagnostics Check	OK	01:52

Data Download Date/Time: 11/21/2017 14:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SINANI,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:04

SUBJECT'S LAST NAME: HILL

DATE/TIME OF TEST: 10/28/2017 03:32:00

Diagnostics Check	OK	03:25
Air Blank	0.000	03:25
Control Test	0.079	03:25
Air Blank	0.000	03:26
Subject Sample # 1	0.197	03:27
Breath Volume	2.632	03:27
Air Blank	0.000	03:27
Air Blank	0.000	03:29
Subject Sample # 2	0.200	03:30
Breath Volume	2.464	03:30
Air Blank	0.000	03:31
Control Test	0.078	03:31
Air Blank	0.000	03:31
Diagnostics Check	OK	03:32

Data Download Date/Time: 11/21/2017 14:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DE JESUS, V
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:16

SUBJECT'S LAST NAME: ANGELES-MONTIEL

DATE/TIME OF TEST: 10/29/2017 00:43:24

Diagnostics Check	OK	00:36
Air Blank	0.000	00:36
Control Test	0.079	00:36
Air Blank	0.000	00:37
Subject Sample # 1	0.197	00:38
Breath Volume	2.531	00:38
Air Blank	0.000	00:39
Air Blank	0.000	00:41
Subject Sample # 2	0.200	00:41
Breath Volume	3.160	00:41
Air Blank	0.000	00:42
Control Test	0.078	00:42
Air Blank	0.000	00:43
Diagnostics Check	OK	00:43

Data Download Date/Time: 11/21/2017 14:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRENNAN, D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:18

SUBJECT'S LAST NAME: KITTEL

DATE/TIME OF TEST: 10/29/2017 01:44:43

Diagnostics Check	OK	01:38
Air Blank	0.000	01:38
Control Test	0.080	01:39
Air Blank	0.000	01:39
Subject Sample # 1	0.109	01:40
Breath Volume	5.089	01:40
Air Blank	0.000	01:40
Air Blank	0.000	01:42
Subject Sample # 2	0.113	01:43
Breath Volume	4.703	01:43
Air Blank	0.000	01:43
Control Test	0.080	01:44
Air Blank	0.000	01:44
Diagnostics Check	OK	01:44

Data Download Date/Time: 11/21/2017 14:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARLAUG,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 10/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:19

SUBJECT'S LAST NAME: JULIEN

DATE/TIME OF TEST: 10/29/2017 02:46:59

Diagnostics Check	OK	02:39
Air Blank	0.000	02:39
Control Test	0.079	02:39
Air Blank	0.000	02:40
Subject Sample # 1	0.177	02:41
Breath Volume	2.000	02:41
Air Blank	0.000	02:42
Air Blank	0.000	02:43
Subject Sample # 2	0.174	02:45
Breath Volume	1.535	02:45
Air Blank	0.000	02:46
Control Test	0.079	02:46
Air Blank	0.000	02:46
Diagnostics Check	OK	02:46

Data Download Date/Time: 11/21/2017 14:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRENNAN,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:42

SUBJECT'S LAST NAME: SODRE

DATE/TIME OF TEST: 10/29/2017 04:09:04

Diagnostics Check	OK	04:02
Air Blank	0.000	04:03
Control Test	0.079	04:03
Air Blank	0.000	04:03
Subject Sample # 1	0.217	04:04
Breath Volume	1.597	04:04
Air Blank	0.000	04:05
Air Blank	0.000	04:06
Subject Sample # 2	0.230	04:07
Breath Volume	2.289	04:07
Air Blank	0.000	04:08
Control Test	0.078	04:08
Air Blank	0.000	04:08
Diagnostics Check	OK	04:09

Data Download Date/Time: 11/21/2017 14:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRIMBLE, TERESA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MACIAS, A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:47

SUBJECT'S LAST NAME: SANDERS

DATE/TIME OF TEST: 10/29/2017 22:14:27

Diagnostics Check	OK	22:08
Air Blank	0.000	22:08
Control Test	0.080	22:08
Air Blank	0.000	22:09
Subject Sample # 1	0.147	22:09
Breath Volume	2.417	22:09
Air Blank	0.000	22:10
Air Blank	0.000	22:12
Subject Sample # 2	0.149	22:12
Breath Volume	2.605	22:12
Air Blank	0.000	22:13
Control Test	0.080	22:13
Air Blank	0.000	22:14
Diagnostics Check	OK	22:14

Data Download Date/Time: 11/21/2017 14:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARLAUG,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 10/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:24

SUBJECT'S LAST NAME: CURRIE

DATE/TIME OF TEST: 10/31/2017 03:51:20

Diagnostics Check	OK	03:44
Air Blank	0.000	03:45
Control Test	0.080	03:45
Air Blank	0.000	03:45
Subject Sample # 1	0.176	03:46
Breath Volume	4.289	03:46
Air Blank	0.000	03:47
Air Blank	0.000	03:49
Subject Sample # 2	0.179	03:49
Breath Volume	4.347	03:49
Air Blank	0.000	03:50
Control Test	0.080	03:50
Air Blank	0.000	03:51
Diagnostics Check	OK	03:51

Data Download Date/Time: 11/21/2017 14:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PECCI,Z,A
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: PECCI,Z
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 09/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31715080A3
DRY GAS EXPIRY DATE: 12/05/2017
VIOLATION CODE: DUI
ARREST TIME: 06:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:50

SUBJECT'S LAST NAME: HANKINS

DATE/TIME OF TEST: 10/01/2017 07:17:06

Diagnostics Check	OK	07:11
Air Blank	0.000	07:11
Control Test	0.077	07:11
Air Blank	0.000	07:12
Subject Sample # 1	0.170	07:12
Breath Volume	2.976	07:12
Air Blank	0.000	07:13
Air Blank	0.000	07:15
Subject Sample # 2	0.165	07:15
Breath Volume	2.695	07:15
Air Blank	0.000	07:16
Control Test	0.078	07:16
Air Blank	0.000	07:17
Diagnostics Check	OK	07:17

Data Download Date/Time: 10/17/2017 19:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PFEIFLE,RYAN,P
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: PFEIFLE,RYAN
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 09/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31715080A3
DRY GAS EXPIRY DATE: 12/05/2017
VIOLATION CODE: DUI
ARREST TIME: 09:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:50

SUBJECT'S LAST NAME: GITTIN

DATE/TIME OF TEST: 10/02/2017 10:24:23

Diagnostics Check	OK	10:18
Air Blank	0.000	10:18
Control Test	0.078	10:19
Air Blank	0.000	10:19
Subject Sample # 1	0.000	10:20
Breath Volume	2.015	10:20
Air Blank	0.000	10:20
Air Blank	0.000	10:22
Subject Sample # 2	0.000	10:23
Breath Volume	2.375	10:23
Air Blank	0.000	10:23
Control Test	0.077	10:23
Air Blank	0.000	10:24
Diagnostics Check	OK	10:24

Data Download Date/Time: 10/17/2017 19:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PATEL,MITHIL,V
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: PATEL,MITHIL
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 09/19/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 31715080A3
DRY GAS EXPIRY DATE: 12/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: BAKER

DATE/TIME OF TEST: 10/03/2017 02:39:15

Diagnostics Check	OK	02:27	
Air Blank	0.000	02:28	
Control Test	0.077	02:28	
Air Blank	0.000	02:29	
Subject Sample	# 1	02:32	No Sample Provided
Breath Volume	0.000	02:32	
Air Blank	0.000	02:32	
Air Blank	0.000	02:34	
Subject Sample	# 2	02:37	No Sample Provided
Breath Volume	0.000	02:37	
Air Blank	0.000	02:38	
Control Test	0.078	02:38	
Air Blank	0.000	02:39	
Diagnostics Check	OK	02:39	

Data Download Date/Time: 10/17/2017 19:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PATEL,MITHIL,V
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: PATEL,MITHIL
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 09/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31715080A3
DRY GAS EXPIRY DATE: 12/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: BAKER

DATE/TIME OF TEST: 10/03/2017 02:49:46

Diagnostics Check	OK	02:43
Air Blank	0.000	02:43
Control Test	0.078	02:44
Air Blank	0.000	02:44
Subject Sample # 1	0.164	02:45
Breath Volume	2.496	02:45
Air Blank	0.000	02:46
Air Blank	0.000	02:47
Subject Sample # 2	0.159	02:48
Breath Volume	1.488	02:48
Air Blank	0.000	02:48
Control Test	0.078	02:49
Air Blank	0.000	02:49
Diagnostics Check	OK	02:49

Data Download Date/Time: 10/17/2017 19:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 09/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31715080A3
DRY GAS EXPIRY DATE: 12/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:46

SUBJECT'S LAST NAME: TROCHEZ

DATE/TIME OF TEST: 10/05/2017 01:28:52

Diagnostics Check	OK	01:17	
Air Blank	0.000	01:17	
Control Test	0.077	01:18	
Air Blank	0.000	01:18	
Subject Sample	# 1	01:21	No Sample Provided
Breath Volume	0.000	01:21	
Air Blank	0.000	01:22	
Air Blank	0.000	01:24	
Subject Sample	# 2	01:27	No Sample Provided
Breath Volume	0.000	01:27	
Air Blank	0.000	01:27	
Control Test	0.078	01:28	
Air Blank	0.000	01:28	
Diagnostics Check	OK	01:28	

Data Download Date/Time: 10/17/2017 19:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HEITFELD,CHRISTOPHER,J
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: HEITFELD,CHRISTOPHER
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 09/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31715080A3
DRY GAS EXPIRY DATE: 12/05/2017
VIOLATION CODE: DUI
ARREST TIME: 12:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:07

SUBJECT'S LAST NAME: TSCHUDI

DATE/TIME OF TEST: 10/05/2017 13:33:42

Diagnostics Check	OK	13:27	
Air Blank	0.000	13:27	
Control Test	0.077	13:28	
Air Blank	0.000	13:28	
Subject Sample	# 1 0.391	13:29	
Breath Volume	2.867	13:29	
Air Blank	0.000	13:29	
Air Blank	0.000	13:31	
Subject Sample	# 2	13:31	Slope Not Met
Breath Volume	1.210	13:31	
Air Blank	0.000	13:32	
Control Test	0.076	13:33	
Air Blank	0.000	13:33	
Diagnostics Check	OK	13:33	

Data Download Date/Time: 10/17/2017 19:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HEITFELD,CHRISTOPHER,J
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: HEITFELD,CHRISTOPHER
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 09/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31715080A3
DRY GAS EXPIRY DATE: 12/05/2017
VIOLATION CODE: DUI
ARREST TIME: 12:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:07

SUBJECT'S LAST NAME: TSCHUDI

DATE/TIME OF TEST: 10/05/2017 13:40:22

Diagnostics Check	OK	13:36	
Air Blank	0.000	13:37	
Control Test	0.077	13:37	
Air Blank	0.000	13:38	
Subject Sample	# 1	13:38	Slope Not Met
Breath Volume	1.851	13:38	
Air Blank	0.000	13:39	
Control Test	0.077	13:39	
Air Blank	0.000	13:40	
Diagnostics Check	OK	13:40	

Data Download Date/Time: 10/17/2017 19:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 09/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31715080A3
DRY GAS EXPIRY DATE: 12/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:05

SUBJECT'S LAST NAME: PEREZ-ROBLERO

DATE/TIME OF TEST: 10/05/2017 21:45:19

Diagnostics Check	OK	21:32	
Air Blank	0.000	21:32	
Control Test	0.078	21:32	
Air Blank	0.000	21:33	
Subject Sample	# 1 0.230	21:35	
Breath Volume	1.175	21:35	
Air Blank	0.000	21:36	
Air Blank	0.000	21:37	
Subject Sample	# 2 0.244	21:40	Volume Not Met
Breath Volume	0.855	21:40	
Air Blank	0.000	21:41	
Air Blank	0.000	21:43	
Subject Sample	# 3 0.216	21:43	
Breath Volume	1.429	21:43	
Air Blank	0.000	21:44	
Control Test	0.074	21:44	Control Outside Tolerance
Air Blank	0.000	21:45	

Data Download Date/Time: 10/17/2017 19:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 09/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31715080A3
DRY GAS EXPIRY DATE: 12/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:05

SUBJECT'S LAST NAME: PEREZ-ROBLERO

DATE/TIME OF TEST: 10/05/2017 22:02:47

Diagnostics Check		OK	21:46	
Air Blank		0.000	21:47	
Control Test		0.076	21:47	
Air Blank		0.000	21:48	
Subject Sample	# 1	0.222	21:51	Volume Not Met
Breath Volume		0.000	21:51	
Air Blank		0.000	21:51	
Air Blank		0.000	21:53	
Subject Sample	# 2	0.234	21:55	
Breath Volume		1.910	21:55	
Air Blank		0.000	21:56	
Air Blank		0.000	21:57	
Subject Sample	# 3	0.239	22:01	Volume Not Met
Breath Volume		0.875	22:01	
Air Blank		0.000	22:01	
Control Test		0.075	22:02	
Air Blank		0.000	22:02	
Diagnostics Check		OK	22:02	

Data Download Date/Time: 10/17/2017 19:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 09/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31715080A3
DRY GAS EXPIRY DATE: 12/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:05

SUBJECT'S LAST NAME: PEREZ-ROBLERO

DATE/TIME OF TEST: 10/05/2017 22:17:38

Diagnostics Check		OK	22:04	
Air Blank		0.000	22:04	
Control Test		0.075	22:05	
Air Blank		0.000	22:05	
Subject Sample	# 1	0.217	22:08	Volume Not Met
Breath Volume		1.089	22:08	
Air Blank		0.000	22:09	
Air Blank		0.000	22:11	
Subject Sample	# 2	0.235	22:13	
Breath Volume		2.332	22:13	
Air Blank		0.000	22:14	
Air Blank		0.000	22:16	
Subject Sample	# 3		22:16	Subject Test Refused
Breath Volume		0.000	22:16	
Air Blank		0.000	22:16	
Control Test		0.075	22:17	
Air Blank		0.000	22:17	
Diagnostics Check		OK	22:17	

Data Download Date/Time: 10/17/2017 19:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANGERVIL,CLAUDY,P
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: ANGERVIL,CLAUDY
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 09/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31715080A3
DRY GAS EXPIRY DATE: 12/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: SELKINEZ

DATE/TIME OF TEST: 10/06/2017 21:20:49

Diagnostics Check	OK	21:12	
Air Blank	0.000	21:12	
Control Test	0.078	21:13	
Air Blank	0.000	21:13	
Subject Sample	# 1 0.243	21:15	
Breath Volume	2.652	21:15	
Air Blank	0.000	21:16	
Air Blank	0.000	21:18	
Subject Sample	# 2	21:19	Subject Test Refused
Breath Volume	0.000	21:19	
Air Blank	0.000	21:19	
Control Test	0.077	21:20	
Air Blank	0.000	21:20	
Diagnostics Check	OK	21:20	

Data Download Date/Time: 10/17/2017 19:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELLAS,DILLON,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: BELLAS,DILLON
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 09/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31715080A3
DRY GAS EXPIRY DATE: 12/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:11

SUBJECT'S LAST NAME: WATERS

DATE/TIME OF TEST: 10/06/2017 22:37:54

Diagnostics Check	OK	22:31
Air Blank	0.000	22:32
Control Test	0.077	22:32
Air Blank	0.000	22:33
Subject Sample # 1	0.166	22:33
Breath Volume	1.218	22:33
Air Blank	0.000	22:34
Air Blank	0.000	22:35
Subject Sample # 2	0.155	22:36
Breath Volume	3.597	22:36
Air Blank	0.000	22:36
Control Test	0.078	22:37
Air Blank	0.000	22:37
Diagnostics Check	OK	22:37

Data Download Date/Time: 10/17/2017 19:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELLAS,DILLON,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: BELLAS,DILLON
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 09/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31715080A3
DRY GAS EXPIRY DATE: 12/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:32

SUBJECT'S LAST NAME: HYMAN

DATE/TIME OF TEST: 10/08/2017 02:59:09

Diagnostics Check	OK	02:52
Air Blank	0.000	02:53
Control Test	0.077	02:53
Air Blank	0.000	02:53
Subject Sample # 1	0.230	02:54
Breath Volume	1.886	02:54
Air Blank	0.000	02:55
Air Blank	0.000	02:56
Subject Sample # 2	0.224	02:57
Breath Volume	1.968	02:57
Air Blank	0.000	02:58
Control Test	0.077	02:58
Air Blank	0.000	02:59
Diagnostics Check	OK	02:59

Data Download Date/Time: 10/17/2017 19:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELLAS,DILLON,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: BELLAS,DILLON
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 09/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31715080A3
DRY GAS EXPIRY DATE: 12/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:14

SUBJECT'S LAST NAME: FERNANDEZ

DATE/TIME OF TEST: 10/08/2017 21:41:56

Diagnostics Check	OK	21:34
Air Blank	0.000	21:34
Control Test	0.079	21:35
Air Blank	0.000	21:35
Subject Sample # 1	0.045	21:37
Breath Volume	1.296	21:37
Air Blank	0.000	21:38
Air Blank	0.000	21:39
Subject Sample # 2	0.043	21:40
Breath Volume	1.906	21:40
Air Blank	0.000	21:41
Control Test	0.075	21:41
Air Blank	0.000	21:41
Diagnostics Check	OK	21:41

Data Download Date/Time: 10/17/2017 19:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANGERVIL,CLAUDY,P
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: ANGERVIL,CLAUDY
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 09/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31715080A3
DRY GAS EXPIRY DATE: 12/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: MARTLAND

DATE/TIME OF TEST: 10/09/2017 00:32:21

Diagnostics Check	OK	00:27	
Air Blank	0.000	00:27	
Control Test	0.077	00:28	
Air Blank	0.000	00:28	
Subject Sample	# 1	00:30	Subject Test Refused
Breath Volume	0.000	00:30	
Air Blank	0.000	00:31	
Control Test	0.078	00:31	
Air Blank	0.000	00:32	
Diagnostics Check	OK	00:32	

Data Download Date/Time: 10/17/2017 19:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 09/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31715080A3
DRY GAS EXPIRY DATE: 12/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:13

SUBJECT'S LAST NAME: TOTTH III

DATE/TIME OF TEST: 10/09/2017 23:41:39

Diagnostics Check	OK	23:34
Air Blank	0.000	23:35
Control Test	0.078	23:35
Air Blank	0.000	23:36
Subject Sample # 1	0.208	23:37
Breath Volume	2.519	23:37
Air Blank	0.000	23:37
Air Blank	0.000	23:39
Subject Sample # 2	0.197	23:40
Breath Volume	2.054	23:40
Air Blank	0.000	23:40
Control Test	0.077	23:41
Air Blank	0.000	23:41
Diagnostics Check	OK	23:41

Data Download Date/Time: 10/17/2017 19:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PECCI,Z,A
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: PECCI,Z
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 09/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31715080A3
DRY GAS EXPIRY DATE: 12/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:19

SUBJECT'S LAST NAME: SHERIFFE

DATE/TIME OF TEST: 10/10/2017 02:52:11

Diagnostics Check	OK	02:45
Air Blank	0.000	02:46
Control Test	0.078	02:46
Air Blank	0.000	02:47
Subject Sample # 1	0.122	02:47
Breath Volume	1.820	02:47
Air Blank	0.000	02:48
Air Blank	0.000	02:50
Subject Sample # 2	0.123	02:50
Breath Volume	2.296	02:50
Air Blank	0.000	02:51
Control Test	0.078	02:51
Air Blank	0.000	02:52
Diagnostics Check	OK	02:52

Data Download Date/Time: 10/17/2017 19:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 09/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31715080A3
DRY GAS EXPIRY DATE: 12/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:31

SUBJECT'S LAST NAME: BONHEUR

DATE/TIME OF TEST: 10/10/2017 23:02:28

Diagnostics Check	OK	22:56
Air Blank	0.000	22:56
Control Test	0.078	22:57
Air Blank	0.000	22:57
Subject Sample # 1	0.013	22:57
Breath Volume	1.335	22:57
Air Blank	0.000	22:58
Air Blank	0.000	23:00
Subject Sample # 2	0.012	23:01
Breath Volume	2.765	23:01
Air Blank	0.000	23:01
Control Test	0.079	23:01
Air Blank	0.000	23:02
Diagnostics Check	OK	23:02

Data Download Date/Time: 10/17/2017 19:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEE,COREY,C
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: LEE,COREY
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 09/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31715080A3
DRY GAS EXPIRY DATE: 12/05/2017
VIOLATION CODE: DUI
ARREST TIME: 17:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:00

SUBJECT'S LAST NAME: LA HERA

DATE/TIME OF TEST: 10/11/2017 18:28:20

Diagnostics Check	OK	18:21
Air Blank	0.000	18:22
Control Test	0.078	18:22
Air Blank	0.000	18:23
Subject Sample # 1	0.171	18:23
Breath Volume	1.695	18:23
Air Blank	0.000	18:24
Air Blank	0.000	18:26
Subject Sample # 2	0.168	18:26
Breath Volume	1.640	18:26
Air Blank	0.000	18:27
Control Test	0.078	18:27
Air Blank	0.000	18:28
Diagnostics Check	OK	18:28

Data Download Date/Time: 10/17/2017 19:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 09/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31715080A3
DRY GAS EXPIRY DATE: 12/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: BEAMS JR

DATE/TIME OF TEST: 10/14/2017 01:40:07

Diagnostics Check	OK	01:34
Air Blank	0.000	01:34
Control Test	0.078	01:34
Air Blank	0.000	01:35
Subject Sample # 1	0.153	01:35
Breath Volume	1.871	01:35
Air Blank	0.000	01:36
Air Blank	0.000	01:38
Subject Sample # 2	0.147	01:38
Breath Volume	2.675	01:38
Air Blank	0.000	01:39
Control Test	0.078	01:39
Air Blank	0.000	01:40
Diagnostics Check	OK	01:40

Data Download Date/Time: 10/17/2017 19:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAYLOR,DEVON,M
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: TAYLOR,DEVON
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 09/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31715080A3
DRY GAS EXPIRY DATE: 12/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: WALTER

DATE/TIME OF TEST: 10/14/2017 02:11:48

Diagnostics Check	OK	02:04
Air Blank	0.000	02:05
Control Test	0.078	02:05
Air Blank	0.000	02:05
Subject Sample # 1	0.234	02:07
Breath Volume	2.351	02:07
Air Blank	0.000	02:07
Air Blank	0.000	02:09
Subject Sample # 2	0.225	02:10
Breath Volume	3.023	02:10
Air Blank	0.000	02:10
Control Test	0.078	02:11
Air Blank	0.000	02:11
Diagnostics Check	OK	02:11

Data Download Date/Time: 10/17/2017 19:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PECCI,Z,A
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: PECCI,Z
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 09/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31715080A3
DRY GAS EXPIRY DATE: 12/05/2017
VIOLATION CODE: DUI
ARREST TIME: 19:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:18

SUBJECT'S LAST NAME: COX

DATE/TIME OF TEST: 10/14/2017 20:44:59

Diagnostics Check	OK	20:38
Air Blank	0.000	20:39
Control Test	0.078	20:39
Air Blank	0.000	20:40
Subject Sample # 1	0.093	20:40
Breath Volume	3.148	20:40
Air Blank	0.000	20:41
Air Blank	0.000	20:43
Subject Sample # 2	0.081	20:43
Breath Volume	1.464	20:43
Air Blank	0.000	20:44
Control Test	0.077	20:44
Air Blank	0.000	20:44
Diagnostics Check	OK	20:44

Data Download Date/Time: 10/17/2017 19:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLASZYK,JACOB,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: BLASZYK,JACOB
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 09/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31715080A3
DRY GAS EXPIRY DATE: 12/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:41

SUBJECT'S LAST NAME: NEDROW

DATE/TIME OF TEST: 10/15/2017 02:08:28

Diagnostics Check	OK	02:02
Air Blank	0.000	02:02
Control Test	0.078	02:03
Air Blank	0.000	02:03
Subject Sample # 1	0.000	02:04
Breath Volume	2.507	02:04
Air Blank	0.000	02:04
Air Blank	0.000	02:06
Subject Sample # 2	0.000	02:07
Breath Volume	2.242	02:07
Air Blank	0.000	02:07
Control Test	0.078	02:07
Air Blank	0.000	02:08
Diagnostics Check	OK	02:08

Data Download Date/Time: 10/17/2017 19:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACOBSON,DAVID,E
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: JACOBSON,DAVID
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 09/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31715080A3
DRY GAS EXPIRY DATE: 12/05/2017
VIOLATION CODE: DUI
ARREST TIME: 15:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:54

SUBJECT'S LAST NAME: DOBERSTEIN

DATE/TIME OF TEST: 10/15/2017 16:32:00

Diagnostics Check	OK	16:25
Air Blank	0.000	16:26
Control Test	0.077	16:26
Air Blank	0.000	16:26
Subject Sample # 1	0.234	16:27
Breath Volume	1.832	16:27
Air Blank	0.000	16:28
Air Blank	0.000	16:29
Subject Sample # 2	0.236	16:30
Breath Volume	1.632	16:30
Air Blank	0.000	16:31
Control Test	0.078	16:31
Air Blank	0.000	16:31
Diagnostics Check	OK	16:32

Data Download Date/Time: 10/17/2017 19:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLASZYK,JACOB,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: BLASZYK,JACOB
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 09/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31715080A3
DRY GAS EXPIRY DATE: 12/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:06

SUBJECT'S LAST NAME: PRICE

DATE/TIME OF TEST: 10/16/2017 03:43:48

Diagnostics Check	OK	03:34
Air Blank	0.000	03:35
Control Test	0.079	03:35
Air Blank	0.000	03:35
Subject Sample # 1	0.095	03:39
Breath Volume	3.464	03:39
Air Blank	0.000	03:39
Air Blank	0.000	03:41
Subject Sample # 2	0.094	03:42
Breath Volume	3.468	03:42
Air Blank	0.000	03:42
Control Test	0.078	03:43
Air Blank	0.000	03:43
Diagnostics Check	OK	03:43

Data Download Date/Time: 10/17/2017 19:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PFEIFLE,RYAN,P
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: PFEIFLE,RYAN
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 09/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31715080A3
DRY GAS EXPIRY DATE: 12/05/2017
VIOLATION CODE: DUI
ARREST TIME: 14:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:26

SUBJECT'S LAST NAME: STRAIT

DATE/TIME OF TEST: 10/17/2017 14:51:14

Diagnostics Check	OK	14:50	
Air Blank		14:50	RFI Detect
Air Blank	0.000	14:51	

Data Download Date/Time: 10/17/2017 19:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PFEIFLE,RYAN,P
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: PFEIFLE,RYAN
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 09/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31715080A3
DRY GAS EXPIRY DATE: 12/05/2017
VIOLATION CODE: DUI
ARREST TIME: 14:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:26

SUBJECT'S LAST NAME: STRAIT

DATE/TIME OF TEST: 10/17/2017 15:02:22

Diagnostics Check	OK	14:54
Air Blank	0.000	14:55
Control Test	0.078	14:55
Air Blank	0.000	14:56
Subject Sample # 1	0.212	14:57
Breath Volume	1.910	14:57
Air Blank	0.000	14:58
Air Blank	0.000	15:00
Subject Sample # 2	0.219	15:00
Breath Volume	2.085	15:00
Air Blank	0.000	15:01
Control Test	0.075	15:01
Air Blank	0.000	15:02
Diagnostics Check	OK	15:02

Data Download Date/Time: 10/17/2017 19:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAYLOR,DEVON,M
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: TAYLOR,DEVON
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 10/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31715080A3
DRY GAS EXPIRY DATE: 12/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:32

SUBJECT'S LAST NAME: WADSWORTH

DATE/TIME OF TEST: 10/18/2017 01:01:49

Diagnostics Check	OK	00:54
Air Blank	0.000	00:54
Control Test	0.079	00:55
Air Blank	0.000	00:55
Subject Sample # 1	0.090	00:56
Breath Volume	1.511	00:56
Air Blank	0.000	00:57
Air Blank	0.000	00:59
Subject Sample # 2	0.081	01:00
Breath Volume	1.410	01:00
Air Blank	0.000	01:00
Control Test	0.076	01:01
Air Blank	0.000	01:01
Diagnostics Check	OK	01:01

Data Download Date/Time: 11/09/2017 14:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 10/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31715080A3
DRY GAS EXPIRY DATE: 12/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:44

SUBJECT'S LAST NAME: WEDGWOOD

DATE/TIME OF TEST: 10/20/2017 02:11:44

Diagnostics Check	OK	02:05
Air Blank	0.000	02:06
Control Test	0.078	02:06
Air Blank	0.000	02:06
Subject Sample # 1	0.175	02:07
Breath Volume	3.484	02:07
Air Blank	0.000	02:07
Air Blank	0.000	02:09
Subject Sample # 2	0.168	02:10
Breath Volume	3.464	02:10
Air Blank	0.000	02:10
Control Test	0.078	02:11
Air Blank	0.000	02:11
Diagnostics Check	OK	02:11

Data Download Date/Time: 11/09/2017 14:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLASZYK,JACOB,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: BLASZYK,JACOB
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 10/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31715080A3
DRY GAS EXPIRY DATE: 12/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:28

SUBJECT'S LAST NAME: ZACARIAS NOGUEIRA

DATE/TIME OF TEST: 10/21/2017 00:10:07

Diagnostics Check	OK	00:03
Air Blank	0.000	00:03
Control Test	0.078	00:04
Air Blank	0.000	00:04
Subject Sample # 1	0.000	00:05
Breath Volume	4.082	00:05
Air Blank	0.000	00:06
Air Blank	0.000	00:08
Subject Sample # 2	0.000	00:08
Breath Volume	4.273	00:08
Air Blank	0.000	00:09
Control Test	0.078	00:09
Air Blank	0.000	00:10
Diagnostics Check	OK	00:10

Data Download Date/Time: 11/09/2017 14:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELLAS,DILLON,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: BELLAS,DILLON
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 10/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31715080A3
DRY GAS EXPIRY DATE: 12/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:38

SUBJECT'S LAST NAME: MICELI

DATE/TIME OF TEST: 10/21/2017 01:07:07

Diagnostics Check		OK	00:58
Air Blank		0.000	00:59
Control Test		0.077	00:59
Air Blank		0.000	01:00
Subject Sample	# 1	0.166	01:01
Breath Volume		2.070	01:01
Air Blank		0.000	01:02
Air Blank		0.000	01:04
Subject Sample	# 2	0.168	01:05
Breath Volume		1.625	01:05
Air Blank		0.000	01:06
Control Test		0.077	01:06
Air Blank		0.000	01:07
Diagnostics Check		OK	01:07

Data Download Date/Time: 11/09/2017 14:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLASZYK,JACOB,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: BLASZYK,JACOB
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 10/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31715080A3
DRY GAS EXPIRY DATE: 12/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:04

SUBJECT'S LAST NAME: MARCOS MARCOS

DATE/TIME OF TEST: 10/23/2017 01:32:18

Diagnostics Check	OK	01:25
Air Blank	0.000	01:25
Control Test	0.078	01:25
Air Blank	0.000	01:26
Subject Sample # 1	0.197	01:27
Breath Volume	2.957	01:27
Air Blank	0.000	01:28
Air Blank	0.000	01:30
Subject Sample # 2	0.204	01:30
Breath Volume	3.339	01:30
Air Blank	0.000	01:31
Control Test	0.077	01:31
Air Blank	0.000	01:32
Diagnostics Check	OK	01:32

Data Download Date/Time: 11/09/2017 14:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PECCI,Z,A
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: PECCI,Z
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 10/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31715080A3
DRY GAS EXPIRY DATE: 12/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:03

SUBJECT'S LAST NAME: ARROYO

DATE/TIME OF TEST: 10/25/2017 01:30:16

Diagnostics Check	OK	01:23
Air Blank	0.000	01:24
Control Test	0.078	01:24
Air Blank	0.000	01:25
Subject Sample # 1	0.130	01:25
Breath Volume	4.679	01:25
Air Blank	0.000	01:26
Air Blank	0.000	01:28
Subject Sample # 2	0.134	01:28
Breath Volume	4.144	01:28
Air Blank	0.000	01:29
Control Test	0.078	01:29
Air Blank	0.000	01:30
Diagnostics Check	OK	01:30

Data Download Date/Time: 11/09/2017 14:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 10/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31715080A3
DRY GAS EXPIRY DATE: 12/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:08

SUBJECT'S LAST NAME: CONNOLLY

DATE/TIME OF TEST: 10/27/2017 22:41:56

Diagnostics Check	OK	22:34
Air Blank	0.000	22:34
Control Test	0.079	22:34
Air Blank	0.000	22:35
Subject Sample # 1	0.157	22:35
Breath Volume	1.207	22:35
Air Blank	0.000	22:36
Air Blank	0.000	22:38
Subject Sample # 2	0.159	22:40
Breath Volume	1.656	22:40
Air Blank	0.000	22:41
Control Test	0.079	22:41
Air Blank	0.000	22:41
Diagnostics Check	OK	22:41

Data Download Date/Time: 11/09/2017 14:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SERRA,CORP,
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: SERRA,CORP
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 10/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31715080A3
DRY GAS EXPIRY DATE: 12/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:14

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 10/28/2017 01:54:27

Diagnostics Check	OK	01:47
Air Blank	0.000	01:48
Control Test	0.081	01:48
Air Blank	0.000	01:49
Subject Sample # 1	0.119	01:49
Breath Volume	2.171	01:49
Air Blank	0.000	01:50
Air Blank	0.000	01:52
Subject Sample # 2	0.119	01:52
Breath Volume	1.671	01:52
Air Blank	0.000	01:53
Control Test	0.077	01:53
Air Blank	0.000	01:54
Diagnostics Check	OK	01:54

Data Download Date/Time: 11/09/2017 14:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PECCI,Z,A
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: PECCI,Z
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 10/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31715080A3
DRY GAS EXPIRY DATE: 12/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: PIERRE

DATE/TIME OF TEST: 10/28/2017 03:38:17

Diagnostics Check	OK	03:32
Air Blank	0.000	03:32
Control Test	0.078	03:32
Air Blank	0.000	03:33
Subject Sample # 1	0.103	03:33
Breath Volume	1.761	03:33
Air Blank	0.000	03:34
Air Blank	0.000	03:36
Subject Sample # 2	0.110	03:36
Breath Volume	2.050	03:36
Air Blank	0.000	03:37
Control Test	0.078	03:37
Air Blank	0.000	03:38
Diagnostics Check	OK	03:38

Data Download Date/Time: 11/09/2017 14:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POLIZZI,CHRISTINE,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: POLIZZI,CHRISTINE
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 10/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31715080A3
DRY GAS EXPIRY DATE: 12/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:29

SUBJECT'S LAST NAME: GORES

DATE/TIME OF TEST: 10/29/2017 02:02:18

Diagnostics Check	OK	01:54
Air Blank	0.000	01:55
Control Test	0.077	01:55
Air Blank	0.000	01:56
Subject Sample # 1	0.134	01:57
Breath Volume	1.183	01:57
Air Blank	0.000	01:57
Air Blank	0.000	01:59
Subject Sample # 2	0.128	02:00
Breath Volume	1.402	02:00
Air Blank	0.000	02:01
Control Test	0.077	02:01
Air Blank	0.000	02:02
Diagnostics Check	OK	02:02

Data Download Date/Time: 11/09/2017 14:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POLIZZI,CHRISTINE,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: POLIZZI,CHRISTINE
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 10/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31715080A3
DRY GAS EXPIRY DATE: 12/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: FISCOR

DATE/TIME OF TEST: 10/30/2017 01:44:36

Diagnostics Check	OK	01:38
Air Blank	0.000	01:38
Control Test	0.078	01:38
Air Blank	0.000	01:39
Subject Sample # 1	0.168	01:40
Breath Volume	1.601	01:40
Air Blank	0.000	01:40
Air Blank	0.000	01:42
Subject Sample # 2	0.159	01:43
Breath Volume	1.375	01:43
Air Blank	0.000	01:43
Control Test	0.078	01:44
Air Blank	0.000	01:44
Diagnostics Check	OK	01:44

Data Download Date/Time: 11/09/2017 14:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRIMBLE, TERESA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER, A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:14

SUBJECT'S LAST NAME: STAFFORD

DATE/TIME OF TEST: 10/01/2017 02:43:26

Diagnostics Check	OK	02:34	
Air Blank	0.000	02:35	
Control Test	0.079	02:35	
Air Blank	0.000	02:36	
Subject Sample	# 1	02:39	No Sample Provided
Breath Volume	0.000	02:39	
Air Blank	0.000	02:39	
Air Blank	0.000	02:41	
Subject Sample	# 2	02:41	Subject Test Refused
Breath Volume	0.000	02:41	
Air Blank	0.000	02:42	
Control Test	0.080	02:42	
Air Blank	0.000	02:43	
Diagnostics Check	OK	02:43	

Data Download Date/Time: 10/04/2017 19:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MACIAS, A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:06

SUBJECT'S LAST NAME: MARCOS MARTINEZ

DATE/TIME OF TEST: 10/01/2017 22:39:45

Diagnostics Check	OK	22:26
Air Blank	0.000	22:27
Control Test	0.080	22:27
Air Blank	0.000	22:28
Subject Sample # 1	0.183	22:30
Breath Volume	3.140	22:30
Air Blank	0.000	22:31
Air Blank	0.000	22:33
Subject Sample # 2	0.155	22:34
Breath Volume	1.519	22:34
Air Blank	0.000	22:34
Air Blank	0.000	22:36
Subject Sample # 3	0.165	22:37
Breath Volume	1.531	22:37
Air Blank	0.000	22:38
Control Test	0.078	22:39
Air Blank	0.000	22:39
Diagnostics Check	OK	22:39

Data Download Date/Time: 10/04/2017 19:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DAWSY, B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:34

SUBJECT'S LAST NAME: HAHN

DATE/TIME OF TEST: 10/02/2017 03:02:21

Diagnostics Check	OK	02:55
Air Blank	0.000	02:56
Control Test	0.080	02:56
Air Blank	0.000	02:57
Subject Sample # 1	0.178	02:57
Breath Volume	1.457	02:57
Air Blank	0.000	02:58
Air Blank	0.000	02:59
Subject Sample # 2	0.179	03:00
Breath Volume	2.003	03:00
Air Blank	0.000	03:01
Control Test	0.079	03:01
Air Blank	0.000	03:02
Diagnostics Check	OK	03:02

Data Download Date/Time: 10/04/2017 19:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PAULINO,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:49

SUBJECT'S LAST NAME: COLLINS JR

DATE/TIME OF TEST: 10/03/2017 04:12:54

Diagnostics Check	OK	04:09	
Air Blank	0.000	04:10	
Control Test	0.080	04:10	
Air Blank	0.000	04:11	
Subject Sample	# 1	04:11	Subject Test Refused
Breath Volume	0.000	04:11	
Air Blank	0.000	04:11	
Control Test	0.080	04:12	
Air Blank	0.000	04:12	
Diagnostics Check	OK	04:12	

Data Download Date/Time: 10/04/2017 19:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRENNAN,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:49

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 10/05/2017 05:17:05

Diagnostics Check	OK	05:09
Air Blank	0.000	05:10
Control Test	0.080	05:10
Air Blank	0.000	05:11
Subject Sample # 1	0.150	05:12
Breath Volume	1.859	05:12
Air Blank	0.000	05:12
Air Blank	0.000	05:14
Subject Sample # 2	0.164	05:15
Breath Volume	4.417	05:15
Air Blank	0.000	05:16
Control Test	0.079	05:16
Air Blank	0.000	05:17
Diagnostics Check	OK	05:17

Data Download Date/Time: 11/21/2017 11:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROOKS, BLAKE, T
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: COWART, GAGE
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 09/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1861309
DRY GAS EXPIRY DATE: 05/19/2018
VIOLATION CODE: DUI
ARREST TIME: 02:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: TYRRELL

DATE/TIME OF TEST: 10/10/2017 03:45:47

Diagnostics Check	OK	03:39
Air Blank	0.000	03:40
Control Test	0.075	03:40
Air Blank	0.000	03:41
Subject Sample # 1	0.000	03:41
Breath Volume	1.312	03:41
Air Blank	0.000	03:41
Air Blank	0.000	03:43
Subject Sample # 2	0.000	03:44
Breath Volume	1.738	03:44
Air Blank	0.000	03:44
Control Test	0.077	03:45
Air Blank	0.000	03:45
Diagnostics Check	OK	03:45

Data Download Date/Time: 10/31/2017 16:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROOKS, BLAKE, T
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: COWART, GAGE
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 09/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1861309
DRY GAS EXPIRY DATE: 05/19/2018
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:52

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 10/11/2017 01:16:05

Diagnostics Check	OK	01:12	
Air Blank	0.000	01:12	
Control Test	0.079	01:13	
Air Blank	0.000	01:13	
Subject Sample	# 1	01:14	Subject Test Refused
Breath Volume	0.000	01:14	
Air Blank	0.000	01:15	
Control Test	0.079	01:15	
Air Blank	0.000	01:15	
Diagnostics Check	OK	01:16	

Data Download Date/Time: 10/31/2017 16:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSS,ROBERT,R
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: MOSS,ROBERT
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 09/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1861309
DRY GAS EXPIRY DATE: 05/19/2018
VIOLATION CODE: DUI
ARREST TIME: 04:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:23

SUBJECT'S LAST NAME: GILMORE JR

DATE/TIME OF TEST: 10/14/2017 05:19:49

Diagnostics Check	OK	05:13
Air Blank	0.000	05:14
Control Test	0.082	05:14
Air Blank	0.000	05:15
Subject Sample # 1	0.000	05:15
Breath Volume	2.410	05:15
Air Blank	0.000	05:16
Air Blank	0.000	05:18
Subject Sample # 2	0.000	05:18
Breath Volume	2.269	05:18
Air Blank	0.000	05:18
Control Test	0.080	05:19
Air Blank	0.000	05:19
Diagnostics Check	OK	05:19

Data Download Date/Time: 10/31/2017 16:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROOKS, BLAKE, T
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: BROOKS, BLAKE
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 09/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1861309
DRY GAS EXPIRY DATE: 05/19/2018
VIOLATION CODE: DUI
ARREST TIME: 23:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:31

SUBJECT'S LAST NAME: VANHOUTEN

DATE/TIME OF TEST: 10/15/2017 00:19:54

Diagnostics Check	OK	00:13
Air Blank	0.000	00:13
Control Test	0.082	00:14
Air Blank	0.000	00:14
Subject Sample # 1	0.093	00:15
Breath Volume	3.605	00:15
Air Blank	0.000	00:15
Air Blank	0.000	00:17
Subject Sample # 2	0.095	00:18
Breath Volume	3.652	00:18
Air Blank	0.000	00:19
Control Test	0.082	00:19
Air Blank	0.000	00:19
Diagnostics Check	OK	00:19

Data Download Date/Time: 10/31/2017 16:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROOKS, BLAKE, T
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: BROOKS, BLAKE
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 09/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1861309
DRY GAS EXPIRY DATE: 05/19/2018
VIOLATION CODE: DUI
ARREST TIME: 00:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:28

SUBJECT'S LAST NAME: PATTON

DATE/TIME OF TEST: 10/16/2017 00:54:44

Diagnostics Check	OK	00:48
Air Blank	0.000	00:48
Control Test	0.082	00:48
Air Blank	0.000	00:49
Subject Sample # 1	0.155	00:50
Breath Volume	4.562	00:50
Air Blank	0.000	00:50
Air Blank	0.000	00:52
Subject Sample # 2	0.157	00:53
Breath Volume	5.070	00:53
Air Blank	0.000	00:53
Control Test	0.080	00:54
Air Blank	0.000	00:54
Diagnostics Check	OK	00:54

Data Download Date/Time: 10/31/2017 16:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROOKS, BLAKE, T
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: BROOKS, BLAKE
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 09/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1861309
DRY GAS EXPIRY DATE: 05/19/2018
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: MANSO

DATE/TIME OF TEST: 10/19/2017 00:50:20

Diagnostics Check	OK	00:44
Air Blank	0.000	00:44
Control Test	0.082	00:44
Air Blank	0.000	00:45
Subject Sample # 1	0.000	00:45
Breath Volume	3.335	00:45
Air Blank	0.000	00:46
Air Blank	0.000	00:48
Subject Sample # 2	0.000	00:48
Breath Volume	3.484	00:48
Air Blank	0.000	00:49
Control Test	0.081	00:49
Air Blank	0.000	00:50
Diagnostics Check	OK	00:50

Data Download Date/Time: 10/31/2017 16:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROOKS, BLAKE, T
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: BROOKS, BLAKE
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 09/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1861309
DRY GAS EXPIRY DATE: 05/19/2018
VIOLATION CODE: DUI
ARREST TIME: 00:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: FYFFE

DATE/TIME OF TEST: 10/28/2017 01:21:44

Diagnostics Check		OK	01:15
Air Blank		0.000	01:15
Control Test		0.081	01:16
Air Blank		0.000	01:16
Subject Sample	# 1	0.155	01:17
Breath Volume		2.238	01:17
Air Blank		0.000	01:18
Air Blank		0.000	01:19
Subject Sample	# 2	0.161	01:20
Breath Volume		2.117	01:20
Air Blank		0.000	01:20
Control Test		0.080	01:21
Air Blank		0.000	01:21
Diagnostics Check		OK	01:21

Data Download Date/Time: 10/31/2017 16:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSS,ROBERT,R
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: MOSS,ROBERT
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 09/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1861309
DRY GAS EXPIRY DATE: 05/19/2018
VIOLATION CODE: DUI
ARREST TIME: 04:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:14

SUBJECT'S LAST NAME: MITCHELL

DATE/TIME OF TEST: 10/29/2017 04:53:22

Diagnostics Check	OK	04:46
Air Blank	0.000	04:46
Control Test	0.082	04:47
Air Blank	0.000	04:47
Subject Sample # 1	0.042	04:48
Breath Volume	1.531	04:48
Air Blank	0.000	04:48
Air Blank	0.000	04:50
Subject Sample # 2	0.037	04:51
Breath Volume	1.308	04:51
Air Blank	0.000	04:52
Control Test	0.081	04:52
Air Blank	0.000	04:53
Diagnostics Check	OK	04:53

Data Download Date/Time: 10/31/2017 16:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: LYNN HAVEN PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RADCLIFF, MICHAEL, C
BREATH TEST OPERATOR AGENCY: Lynn Haven PD
ARRESTING OFFICER: BOHRER, MARK
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 10/18/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:44

SUBJECT'S LAST NAME: BENEVENTO

DATE/TIME OF TEST: 10/24/2017 00:20:22

Diagnostics Check	OK	00:19	
Air Blank	0.000	00:19	
Control Test	0.000	00:19	RFI Detect
Air Blank	0.000	00:20	

Data Download Date/Time: 11/30/2017 03:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: LYNN HAVEN PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RADCLIFF, MICHAEL, C
BREATH TEST OPERATOR AGENCY: Lynn Haven PD
ARRESTING OFFICER: BOHRER, MARK
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 10/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:44

SUBJECT'S LAST NAME: BENEVENTO

DATE/TIME OF TEST: 10/24/2017 00:30:17

Diagnostics Check	OK	00:22
Air Blank	0.000	00:23
Control Test	0.075	00:23
Air Blank	0.000	00:24
Subject Sample # 1	0.138	00:25
Breath Volume	1.585	00:25
Air Blank	0.000	00:26
Air Blank	0.000	00:28
Subject Sample # 2	0.126	00:28
Breath Volume	1.921	00:28
Air Blank	0.000	00:29
Control Test	0.076	00:29
Air Blank	0.000	00:30
Diagnostics Check	OK	00:30

Data Download Date/Time: 11/30/2017 03:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: LYNN HAVEN PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RADCLIFF, MICHAEL, C
BREATH TEST OPERATOR AGENCY: Lynn Haven PD
ARRESTING OFFICER: SOUTH, DAMIAN
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 10/18/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:59

SUBJECT'S LAST NAME: WEBSTER

DATE/TIME OF TEST: 10/25/2017 00:35:28

Diagnostics Check	OK	00:28
Air Blank	0.000	00:28
Control Test	0.078	00:29
Air Blank	0.000	00:29
Subject Sample # 1	0.365	00:30
Breath Volume	1.632	00:30
Air Blank	0.000	00:30
Air Blank	0.000	00:32
Subject Sample # 2	0.355	00:33
Breath Volume	1.644	00:33
Air Blank	0.000	00:34
Control Test	0.078	00:34
Air Blank	0.000	00:35
Diagnostics Check	OK	00:35

Data Download Date/Time: 11/30/2017 03:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: LYNN HAVEN PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUSCHMANN,BRIAN,S
BREATH TEST OPERATOR AGENCY: Lynn Haven PD
ARRESTING OFFICER: BUSCHMANN,BRIAN
ARRESTING OFFICER AGENCY: Lynn Haven PD
LAST AGENCY INSPECTION: 10/18/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:40

SUBJECT'S LAST NAME: VALLE

DATE/TIME OF TEST: 10/29/2017 22:11:18

Diagnostics Check	OK	22:08	
Air Blank	0.000	22:08	
Control Test	0.078	22:09	
Air Blank	0.000	22:09	
Subject Sample	# 1	22:09	Subject Test Refused
Breath Volume	0.000	22:09	
Air Blank	0.000	22:10	
Control Test	0.077	22:10	
Air Blank	0.000	22:11	
Diagnostics Check	OK	22:11	

Data Download Date/Time: 11/30/2017 03:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: LYNN HAVEN PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STREICHERT,MICHAEL,L
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: STREICHERT,MICHAEL
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 10/18/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:28

SUBJECT'S LAST NAME: LISEBY

DATE/TIME OF TEST: 10/31/2017 22:05:14

Diagnostics Check		OK	21:59
Air Blank		0.000	21:59
Control Test		0.077	22:00
Air Blank		0.000	22:00
Subject Sample	# 1	0.240	22:00
Breath Volume		2.964	22:00
Air Blank		0.000	22:01
Air Blank		0.000	22:03
Subject Sample	# 2	0.260	22:03
Breath Volume		2.203	22:03
Air Blank		0.000	22:04
Control Test		0.077	22:04
Air Blank		0.000	22:05
Diagnostics Check		OK	22:05

Data Download Date/Time: 11/30/2017 03:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TUCKER,HAILEY,E
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: DEBERARDINIS,DREW
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 09/20/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:20

SUBJECT'S LAST NAME: EGER

DATE/TIME OF TEST: 10/03/2017 20:48:06

Diagnostics Check	OK	20:41
Air Blank	0.000	20:42
Control Test	0.081	20:42
Air Blank	0.000	20:42
Subject Sample # 1	0.014	20:43
Breath Volume	2.394	20:43
Air Blank	0.000	20:44
Air Blank	0.000	20:46
Subject Sample # 2	0.012	20:46
Breath Volume	2.199	20:46
Air Blank	0.000	20:47
Control Test	0.080	20:47
Air Blank	0.000	20:48
Diagnostics Check	OK	20:48

Data Download Date/Time: 10/18/2017 09:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,ALFONSO,
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: RODRIGUEZ,ALFONSO
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 09/20/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:46

SUBJECT'S LAST NAME: GAGNON

DATE/TIME OF TEST: 10/05/2017 03:20:44

Diagnostics Check	OK	03:12
Air Blank	0.000	03:13
Control Test	0.081	03:13
Air Blank	0.000	03:13
Subject Sample # 1	0.164	03:15
Breath Volume	2.273	03:15
Air Blank	0.000	03:16
Air Blank	0.000	03:17
Subject Sample # 2	0.163	03:19
Breath Volume	2.078	03:19
Air Blank	0.000	03:19
Control Test	0.080	03:20
Air Blank	0.000	03:20
Diagnostics Check	OK	03:20

Data Download Date/Time: 10/18/2017 09:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSON,SARAH,M
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: JOHNSON,SARAH
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 09/20/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:54

SUBJECT'S LAST NAME: BLAIR

DATE/TIME OF TEST: 10/05/2017 03:27:38

Diagnostics Check	OK	03:24	
Air Blank	0.000	03:24	
Control Test	0.081	03:25	
Air Blank	0.000	03:25	
Subject Sample	# 1	03:26	Subject Test Refused
Breath Volume	0.000	03:26	
Air Blank	0.000	03:26	
Control Test	0.080	03:27	
Air Blank	0.000	03:27	
Diagnostics Check	OK	03:27	

Data Download Date/Time: 10/18/2017 09:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,DANNY,W
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: BUSCH,CAROLINE
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 09/20/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: RAMOS

DATE/TIME OF TEST: 10/07/2017 23:52:13

Diagnostics Check	OK	23:45
Air Blank	0.000	23:45
Control Test	0.081	23:46
Air Blank	0.000	23:46
Subject Sample # 1	0.286	23:47
Breath Volume	2.078	23:47
Air Blank	0.000	23:48
Air Blank	0.000	23:50
Subject Sample # 2	0.294	23:50
Breath Volume	2.785	23:50
Air Blank	0.000	23:51
Control Test	0.081	23:51
Air Blank	0.000	23:52
Diagnostics Check	OK	23:52

Data Download Date/Time: 10/18/2017 09:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUSH,DANIEL,C
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: SERWY,C
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 09/20/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:47

SUBJECT'S LAST NAME: HOLM

DATE/TIME OF TEST: 10/13/2017 23:24:46

Diagnostics Check	OK	23:18
Air Blank	0.000	23:18
Control Test	0.082	23:19
Air Blank	0.000	23:19
Subject Sample # 1	0.184	23:20
Breath Volume	2.355	23:20
Air Blank	0.000	23:20
Air Blank	0.000	23:22
Subject Sample # 2	0.197	23:23
Breath Volume	2.429	23:23
Air Blank	0.000	23:23
Control Test	0.081	23:24
Air Blank	0.000	23:24
Diagnostics Check	OK	23:24

Data Download Date/Time: 10/18/2017 09:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,DANNY,W
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: GRAY,DANNY
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 09/20/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: OTUL-CUMMINGS

DATE/TIME OF TEST: 10/14/2017 00:13:56

Diagnostics Check	OK	00:11	
Air Blank	0.000	00:11	
Control Test	0.081	00:11	
Air Blank	0.000	00:12	
Subject Sample	# 1	00:12	Subject Test Refused
Breath Volume	0.000	00:12	
Air Blank	0.000	00:13	
Control Test	0.081	00:13	
Air Blank	0.000	00:13	
Diagnostics Check	OK	00:13	

Data Download Date/Time: 10/18/2017 09:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUSKEVICH,CURT,E
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: GINSBERG,J
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 09/20/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 18:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:15

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 10/15/2017 19:43:38

Diagnostics Check	OK	19:37
Air Blank	0.000	19:38
Control Test	0.081	19:38
Air Blank	0.000	19:38
Subject Sample # 1	0.201	19:39
Breath Volume	1.343	19:39
Air Blank	0.000	19:39
Air Blank	0.000	19:41
Subject Sample # 2	0.204	19:42
Breath Volume	1.468	19:42
Air Blank	0.000	19:42
Control Test	0.081	19:43
Air Blank	0.000	19:43
Diagnostics Check	OK	19:43

Data Download Date/Time: 10/18/2017 09:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KARAU,J.,L
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: HOGAN,W.
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 10/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 14:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:11

SUBJECT'S LAST NAME: MAXWELL

DATE/TIME OF TEST: 10/18/2017 16:38:22

Diagnostics Check	OK	16:31
Air Blank	0.000	16:32
Control Test	0.080	16:32
Air Blank	0.000	16:33
Subject Sample # 1	0.215	16:33
Breath Volume	6.046	16:33
Air Blank	0.000	16:34
Air Blank	0.000	16:36
Subject Sample # 2	0.207	16:36
Breath Volume	3.871	16:36
Air Blank	0.000	16:37
Control Test	0.080	16:37
Air Blank	0.000	16:38
Diagnostics Check	OK	16:38

Data Download Date/Time: 11/08/2017 10:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TOBLER,AMANDA,R
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: TOBLER,AMANDA
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 10/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:05

SUBJECT'S LAST NAME: MONTEJO

DATE/TIME OF TEST: 10/18/2017 23:31:36

Diagnostics Check	OK	23:25
Air Blank	0.000	23:26
Control Test	0.081	23:26
Air Blank	0.000	23:27
Subject Sample # 1	0.000	23:27
Breath Volume	2.871	23:27
Air Blank	0.000	23:27
Air Blank	0.000	23:29
Subject Sample # 2	0.000	23:30
Breath Volume	3.894	23:30
Air Blank	0.000	23:30
Control Test	0.081	23:31
Air Blank	0.000	23:31
Diagnostics Check	OK	23:31

Data Download Date/Time: 11/08/2017 10:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARKLEY,JEFFREY,G
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: BARKLEY,JEFFREY
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 10/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:52

SUBJECT'S LAST NAME: FOLGER

DATE/TIME OF TEST: 10/21/2017 01:18:31

Diagnostics Check	OK	01:12
Air Blank	0.000	01:12
Control Test	0.081	01:12
Air Blank	0.000	01:13
Subject Sample # 1	0.119	01:14
Breath Volume	1.164	01:14
Air Blank	0.000	01:14
Air Blank	0.000	01:16
Subject Sample # 2	0.116	01:17
Breath Volume	2.511	01:17
Air Blank	0.000	01:17
Control Test	0.080	01:17
Air Blank	0.000	01:18
Diagnostics Check	OK	01:18

Data Download Date/Time: 11/08/2017 10:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUSKEVICH,CURT,E
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: GALLO,A
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 10/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:03

SUBJECT'S LAST NAME: WARD

DATE/TIME OF TEST: 10/27/2017 23:36:50

Diagnostics Check	OK	23:30
Air Blank	0.000	23:30
Control Test	0.081	23:31
Air Blank	0.000	23:31
Subject Sample # 1	0.000	23:32
Breath Volume	4.417	23:32
Air Blank	0.000	23:32
Air Blank	0.000	23:34
Subject Sample # 2	0.000	23:35
Breath Volume	3.988	23:35
Air Blank	0.000	23:35
Control Test	0.082	23:36
Air Blank	0.000	23:36
Diagnostics Check	OK	23:36

Data Download Date/Time: 11/08/2017 10:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUSH,DANIEL,C
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: LOOS,S
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 10/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: FREEMAN

DATE/TIME OF TEST: 10/28/2017 01:05:10

Diagnostics Check	OK	00:58
Air Blank	0.000	00:59
Control Test	0.081	00:59
Air Blank	0.000	01:00
Subject Sample # 1	0.069	01:00
Breath Volume	5.082	01:00
Air Blank	0.000	01:01
Air Blank	0.000	01:03
Subject Sample # 2	0.075	01:03
Breath Volume	4.855	01:03
Air Blank	0.000	01:04
Control Test	0.081	01:04
Air Blank	0.000	01:05
Diagnostics Check	OK	01:05

Data Download Date/Time: 11/08/2017 10:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUSH,DANIEL,C
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: CINTRON,M
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 10/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:54

SUBJECT'S LAST NAME: PARADA CASTRO

DATE/TIME OF TEST: 10/28/2017 03:23:54

Diagnostics Check	OK	03:16
Air Blank	0.000	03:16
Control Test	0.081	03:17
Air Blank	0.000	03:17
Subject Sample # 1	0.163	03:19
Breath Volume	2.847	03:19
Air Blank	0.000	03:19
Air Blank	0.000	03:21
Subject Sample # 2	0.162	03:22
Breath Volume	2.160	03:22
Air Blank	0.000	03:23
Control Test	0.080	03:23
Air Blank	0.000	03:23
Diagnostics Check	OK	03:23

Data Download Date/Time: 11/08/2017 10:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARSON,JAMES,A
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: CARSON,JAMES
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 10/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:57

SUBJECT'S LAST NAME: LABRIOLA

DATE/TIME OF TEST: 10/29/2017 00:27:20

Diagnostics Check	OK	00:20
Air Blank	0.000	00:20
Control Test	0.081	00:21
Air Blank	0.000	00:21
Subject Sample # 1	0.175	00:22
Breath Volume	1.464	00:22
Air Blank	0.000	00:23
Air Blank	0.000	00:25
Subject Sample # 2	0.174	00:25
Breath Volume	1.503	00:25
Air Blank	0.000	00:26
Control Test	0.081	00:26
Air Blank	0.000	00:27
Diagnostics Check	OK	00:27

Data Download Date/Time: 11/08/2017 10:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TOBLER,AMANDA,R
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: TOBLER,AMANDA
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 10/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:21

SUBJECT'S LAST NAME: JIMENEZ

DATE/TIME OF TEST: 10/29/2017 02:59:06

Diagnostics Check		OK	02:47	
Air Blank		0.000	02:48	
Control Test		0.082	02:48	
Air Blank		0.000	02:48	
Subject Sample	# 1	0.148	02:52	Volume Not Met
Breath Volume		0.792	02:52	
Air Blank		0.000	02:52	
Air Blank		0.000	02:54	
Subject Sample	# 2	0.165	02:57	Volume Not Met
Breath Volume		0.988	02:57	
Air Blank		0.000	02:58	
Control Test		0.080	02:58	
Air Blank		0.000	02:59	
Diagnostics Check		OK	02:59	

Data Download Date/Time: 11/08/2017 10:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TOBLER,AMANDA,R
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: SCARLATO,FRANK
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 10/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: RAMIREZ

DATE/TIME OF TEST: 10/29/2017 03:16:04

Diagnostics Check	OK	03:08
Air Blank	0.000	03:09
Control Test	0.081	03:09
Air Blank	0.000	03:09
Subject Sample # 1	0.214	03:11
Breath Volume	1.710	03:11
Air Blank	0.000	03:12
Air Blank	0.000	03:13
Subject Sample # 2	0.210	03:14
Breath Volume	1.296	03:14
Air Blank	0.000	03:15
Control Test	0.080	03:15
Air Blank	0.000	03:15
Diagnostics Check	OK	03:16

Data Download Date/Time: 11/08/2017 10:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TOBLER,AMANDA,R
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: TOBLER,AMANDA
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 10/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: MILBURN

DATE/TIME OF TEST: 10/30/2017 03:24:55

Diagnostics Check	OK	03:21	
Air Blank	0.000	03:22	
Control Test	0.082	03:22	
Air Blank	0.000	03:23	
Subject Sample	# 1	03:23	Subject Test Refused
Breath Volume	0.000	03:23	
Air Blank	0.000	03:24	
Control Test	0.081	03:24	
Air Blank	0.000	03:24	
Diagnostics Check	OK	03:24	

Data Download Date/Time: 11/08/2017 10:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUEHLER,DEP,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BUEHLER,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 09/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: BRUNEMAN

DATE/TIME OF TEST: 10/07/2017 22:34:11

Diagnostics Check	OK	22:27
Air Blank	0.000	22:28
Control Test	0.080	22:28
Air Blank	0.000	22:29
Subject Sample # 1	0.261	22:29
Breath Volume	2.503	22:29
Air Blank	0.000	22:30
Air Blank	0.000	22:32
Subject Sample # 2	0.268	22:32
Breath Volume	2.992	22:32
Air Blank	0.000	22:33
Control Test	0.079	22:33
Air Blank	0.000	22:34
Diagnostics Check	OK	22:34

Data Download Date/Time: 10/25/2017 07:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUEHLER,DEP,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BUEHLER,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 09/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:14

SUBJECT'S LAST NAME: GARTH-ZAMORA

DATE/TIME OF TEST: 10/08/2017 03:43:21

Diagnostics Check	OK	03:42	
Air Blank		03:42	RFI Detect
Air Blank	0.000	03:43	

Data Download Date/Time: 10/25/2017 07:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUEHLER,DEP,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BUEHLER,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 09/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:14

SUBJECT'S LAST NAME: GARTH-ZAMORA

DATE/TIME OF TEST: 10/08/2017 03:51:41

Diagnostics Check	OK	03:44
Air Blank	0.000	03:45
Control Test	0.079	03:45
Air Blank	0.000	03:46
Subject Sample # 1	0.227	03:47
Breath Volume	3.726	03:47
Air Blank	0.000	03:47
Air Blank	0.000	03:49
Subject Sample # 2	0.220	03:50
Breath Volume	1.976	03:50
Air Blank	0.000	03:50
Control Test	0.079	03:51
Air Blank	0.000	03:51
Diagnostics Check	OK	03:51

Data Download Date/Time: 10/25/2017 07:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUEHLER,DEP,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BUEHLER,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 09/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:33

SUBJECT'S LAST NAME: LEACH

DATE/TIME OF TEST: 10/20/2017 02:26:02

Diagnostics Check	OK	02:14	
Air Blank	0.000	02:15	
Control Test	0.080	02:15	
Air Blank	0.000	02:15	
Subject Sample	# 1	02:19	No Sample Provided
Breath Volume	0.000	02:19	
Air Blank	0.000	02:19	
Air Blank	0.000	02:21	
Subject Sample	# 2	02:24	No Sample Provided
Breath Volume	0.000	02:24	
Air Blank	0.000	02:25	
Control Test	0.080	02:25	
Air Blank	0.000	02:25	
Diagnostics Check	OK	02:26	

Data Download Date/Time: 10/25/2017 07:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUEHLER,DEP,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BUEHLER,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 09/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:21

SUBJECT'S LAST NAME: FRIELDS

DATE/TIME OF TEST: 10/21/2017 21:51:52

Diagnostics Check	OK	21:42	
Air Blank	0.000	21:42	
Control Test	0.081	21:43	
Air Blank	0.000	21:43	
Subject Sample	# 1 0.169	21:46	
Breath Volume	2.289	21:46	
Air Blank	0.000	21:47	
Air Blank	0.000	21:49	
Subject Sample	# 2	21:50	RFI Detect
Breath Volume	0.000	21:50	
Air Blank	0.000	21:50	
Control Test	0.080	21:51	
Air Blank	0.000	21:51	
Diagnostics Check	OK	21:51	

Data Download Date/Time: 10/25/2017 07:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUEHLER,DEP,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BUEHLER,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 09/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:21

SUBJECT'S LAST NAME: FRIELDS

DATE/TIME OF TEST: 10/21/2017 21:59:28

Diagnostics Check	OK	21:54	
Air Blank	0.000	21:54	
Control Test	0.080	21:55	
Air Blank	0.000	21:55	
Subject Sample # 1	0.157	21:57	
Breath Volume	1.359	21:57	
Air Blank		21:58	RFI Detect
Air Blank	0.000	21:58	
Control Test	0.081	21:58	
Air Blank	0.000	21:59	
Diagnostics Check	OK	21:59	

Data Download Date/Time: 10/25/2017 07:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000862

INSTRUMENT REGISTERED TO: BLOUNTSTOWN PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHIPMAN,KAREN,D
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: SHIPMAN,KAREN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 10/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 12:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: WARD

DATE/TIME OF TEST: 10/06/2017 02:02:51

Diagnostics Check	OK	02:01	
Air Blank	0.000	02:02	
Control Test	0.078	02:02	RFI Detect
Air Blank	0.000	02:02	

Data Download Date/Time: 10/20/2017 11:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000862

INSTRUMENT REGISTERED TO: BLOUNTSTOWN PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHIPMAN,KAREN,D
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: SHIPMAN,KAREN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 10/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 12:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: WARD

DATE/TIME OF TEST: 10/06/2017 02:16:20

Diagnostics Check		OK	02:04	
Air Blank		0.000	02:05	
Control Test		0.079	02:05	
Air Blank		0.000	02:06	
Subject Sample	# 1	0.000	02:09	Volume Not Met
Breath Volume		0.234	02:09	
Air Blank		0.000	02:09	
Air Blank		0.000	02:11	
Subject Sample	# 2	0.000	02:14	Volume Not Met
Breath Volume		0.218	02:14	
Air Blank		0.000	02:15	
Control Test		0.079	02:15	
Air Blank		0.000	02:16	
Diagnostics Check		OK	02:16	

Data Download Date/Time: 10/20/2017 11:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KORZEN,GREG,B
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: KORZEN,GREG
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 09/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 03:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:03

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 10/01/2017 04:45:40

Diagnostics Check	OK	04:42	
Air Blank	0.000	04:42	
Control Test	0.078	04:43	
Air Blank	0.000	04:43	
Subject Sample	# 1	04:44	RFI Detect
Breath Volume	0.039	04:44	
Air Blank	0.000	04:44	
Control Test	0.077	04:45	
Air Blank	0.000	04:45	
Diagnostics Check	OK	04:45	

Data Download Date/Time: 10/01/2017 15:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KORZEN,GREG,B
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: KORZEN,GREG
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 09/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 03:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:03

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 10/01/2017 04:55:13

Diagnostics Check	OK	04:47
Air Blank	0.000	04:47
Control Test	0.078	04:47
Air Blank	0.000	04:48
Subject Sample # 1	0.139	04:50
Breath Volume	1.441	04:50
Air Blank	0.000	04:50
Air Blank	0.000	04:52
Subject Sample # 2	0.146	04:53
Breath Volume	1.464	04:53
Air Blank	0.000	04:54
Control Test	0.077	04:54
Air Blank	0.000	04:55
Diagnostics Check	OK	04:55

Data Download Date/Time: 10/01/2017 15:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOPP,AUSTIN,S
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: JONES,SHANNON
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 10/01/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 21:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:42

SUBJECT'S LAST NAME: SCHAFFER

DATE/TIME OF TEST: 10/03/2017 22:22:55

Diagnostics Check	OK	22:19	
Air Blank	0.000	22:20	
Control Test	0.078	22:20	
Air Blank	0.000	22:21	
Subject Sample	# 1	22:21	Subject Test Refused
Breath Volume	0.000	22:21	
Air Blank	0.000	22:22	
Control Test	0.078	22:22	
Air Blank	0.000	22:22	
Diagnostics Check	OK	22:22	

Data Download Date/Time: 11/16/2017 14:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KEOHANE,KUNIKO,
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: KEOHANE,KUNIKO
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 10/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 10:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:05

SUBJECT'S LAST NAME: HAKE

DATE/TIME OF TEST: 10/04/2017 10:30:04

Diagnostics Check	OK	10:26	
Air Blank	0.000	10:27	
Control Test	0.078	10:27	
Air Blank	0.000	10:28	
Subject Sample # 1	0.000	10:28	
Breath Volume	2.671	10:28	
Air Blank		10:28	Improper Sample
Air Blank	0.000	10:29	
Control Test	0.079	10:29	
Air Blank	0.000	10:29	
Diagnostics Check	OK	10:30	

Data Download Date/Time: 11/16/2017 14:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KEOHANE,KUNIKO,
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: KEOHANE,KUNIKO
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 10/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 10:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:05

SUBJECT'S LAST NAME: HAKE

DATE/TIME OF TEST: 10/04/2017 10:37:53

Diagnostics Check	OK	10:31
Air Blank	0.000	10:32
Control Test	0.079	10:32
Air Blank	0.000	10:33
Subject Sample # 1	0.000	10:33
Breath Volume	2.566	10:33
Air Blank	0.000	10:34
Air Blank	0.000	10:36
Subject Sample # 2	0.000	10:36
Breath Volume	2.226	10:36
Air Blank	0.000	10:36
Control Test	0.079	10:37
Air Blank	0.000	10:37
Diagnostics Check	OK	10:37

Data Download Date/Time: 11/16/2017 14:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,JESSE,A
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: YOUNG,JESSE
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 10/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 04:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: KNOOF

DATE/TIME OF TEST: 10/12/2017 05:10:55

Diagnostics Check	OK	04:59	
Air Blank	0.000	04:59	
Control Test	0.079	05:00	
Air Blank	0.000	05:00	
Subject Sample	# 1 0.117	05:03	Slope Not Level
Breath Volume	1.199	05:03	
Air Blank	0.000	05:04	
Air Blank	0.000	05:06	
Subject Sample	# 2 0.115	05:09	Volume Not Met
Breath Volume	0.625	05:09	
Air Blank	0.000	05:10	
Control Test	0.078	05:10	
Air Blank	0.000	05:10	
Diagnostics Check	OK	05:10	

Data Download Date/Time: 11/16/2017 14:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STOCKTON,JOSEPH,A
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: CONATY,JAY
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 10/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 20:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:01

SUBJECT'S LAST NAME: YODER

DATE/TIME OF TEST: 10/13/2017 21:31:08

Diagnostics Check	OK	21:24
Air Blank	0.000	21:25
Control Test	0.079	21:25
Air Blank	0.000	21:25
Subject Sample # 1	0.269	21:26
Breath Volume	2.074	21:26
Air Blank	0.000	21:27
Air Blank	0.000	21:28
Subject Sample # 2	0.253	21:29
Breath Volume	1.832	21:29
Air Blank	0.000	21:30
Control Test	0.078	21:30
Air Blank	0.000	21:31
Diagnostics Check	OK	21:31

Data Download Date/Time: 11/16/2017 14:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,ALEX,
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: CONATY,JAY
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 10/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 04:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:49

SUBJECT'S LAST NAME: YERO CANETE

DATE/TIME OF TEST: 10/15/2017 07:17:29

Diagnostics Check		OK	07:11
Air Blank		0.000	07:11
Control Test		0.079	07:12
Air Blank		0.000	07:12
Subject Sample	# 1	0.213	07:13
Breath Volume		2.066	07:13
Air Blank		0.000	07:13
Air Blank		0.000	07:15
Subject Sample	# 2	0.196	07:15
Breath Volume		1.593	07:15
Air Blank		0.000	07:16
Control Test		0.077	07:16
Air Blank		0.000	07:17
Diagnostics Check		OK	07:17

Data Download Date/Time: 11/16/2017 14:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELEJALDE RAMIREZ,LINA,M
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: MCGEE,JOHN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 10/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 07:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:11

SUBJECT'S LAST NAME: ELEJALDE RAMIREZ

DATE/TIME OF TEST: 10/16/2017 08:35:55

Diagnostics Check	OK	08:33	
Air Blank	0.000	08:33	
Control Test	0.079	08:33	
Air Blank	0.000	08:34	
Subject Sample	# 1	08:34	Subject Test Refused
Breath Volume	0.000	08:34	
Air Blank	0.000	08:35	
Control Test	0.078	08:35	
Air Blank	0.000	08:35	
Diagnostics Check	OK	08:35	

Data Download Date/Time: 11/16/2017 14:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,JESSE,A
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: YOUNG,JESSE
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 10/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: HARDEN

DATE/TIME OF TEST: 10/21/2017 00:04:08

Diagnostics Check	OK	23:57
Air Blank	0.000	23:57
Control Test	0.078	23:57
Air Blank	0.000	23:58
Subject Sample # 2	0.242	23:59
Breath Volume	1.394	23:59
Air Blank	0.000	00:00
Air Blank	0.000	00:02
Subject Sample # 1	0.225	00:02
Breath Volume	3.136	00:02
Air Blank	0.000	00:03
Control Test	0.077	00:03
Air Blank	0.000	00:04
Diagnostics Check	OK	00:04

Data Download Date/Time: 11/16/2017 14:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOPP,AUSTIN,S
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: HOPP,AUSTIN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 10/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 00:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: PINEDA

DATE/TIME OF TEST: 10/21/2017 01:39:07

Diagnostics Check	OK	01:31
Air Blank	0.000	01:31
Control Test	0.078	01:31
Air Blank	0.000	01:32
Subject Sample # 1	0.189	01:34
Breath Volume	1.343	01:34
Air Blank	0.000	01:35
Air Blank	0.000	01:37
Subject Sample # 2	0.192	01:37
Breath Volume	1.117	01:37
Air Blank	0.000	01:38
Control Test	0.079	01:38
Air Blank	0.000	01:39
Diagnostics Check	OK	01:39

Data Download Date/Time: 11/16/2017 14:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELSMORE,JUSTIN,R
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: MOBLEY,BRITTANY
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 10/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 20:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: WILDER

DATE/TIME OF TEST: 10/23/2017 23:07:09

Diagnostics Check	OK	23:00
Air Blank	0.000	23:01
Control Test	0.079	23:01
Air Blank	0.000	23:02
Subject Sample # 1	0.214	23:02
Breath Volume	2.003	23:02
Air Blank	0.000	23:03
Air Blank	0.000	23:04
Subject Sample # 2	0.214	23:05
Breath Volume	1.832	23:05
Air Blank	0.000	23:06
Control Test	0.079	23:06
Air Blank	0.000	23:07
Diagnostics Check	OK	23:07

Data Download Date/Time: 11/16/2017 14:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOPP,AUSTIN,S
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: JONES,SHANNON
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 10/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 01:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:46

SUBJECT'S LAST NAME: BRAY

DATE/TIME OF TEST: 10/26/2017 03:10:15

Diagnostics Check	OK	03:06	
Air Blank	0.000	03:07	
Control Test	0.079	03:07	
Air Blank	0.000	03:08	
Subject Sample	# 1	03:08	Subject Test Refused
Breath Volume	0.000	03:08	
Air Blank	0.000	03:09	
Control Test	0.079	03:09	
Air Blank	0.000	03:10	
Diagnostics Check	OK	03:10	

Data Download Date/Time: 11/16/2017 14:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,ALEX,
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: RODRIGUEZ,ALEX
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 10/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 16:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:41

SUBJECT'S LAST NAME: MAYNARD

DATE/TIME OF TEST: 10/27/2017 17:10:03

Diagnostics Check	OK	17:03
Air Blank	0.000	17:04
Control Test	0.080	17:04
Air Blank	0.000	17:05
Subject Sample # 1	0.150	17:05
Breath Volume	2.246	17:05
Air Blank	0.000	17:06
Air Blank	0.000	17:08
Subject Sample # 2	0.145	17:08
Breath Volume	1.593	17:08
Air Blank	0.000	17:09
Control Test	0.078	17:09
Air Blank	0.000	17:09
Diagnostics Check	OK	17:10

Data Download Date/Time: 11/16/2017 14:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELSMORE,JUSTIN,R
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: CONATY,JAY
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 10/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 02:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:02

SUBJECT'S LAST NAME: AGUIRRE-TREMINIO

DATE/TIME OF TEST: 10/30/2017 03:29:15

Diagnostics Check		OK	03:23
Air Blank		0.000	03:23
Control Test		0.079	03:23
Air Blank		0.000	03:24
Subject Sample	# 1	0.179	03:24
Breath Volume		2.812	03:24
Air Blank		0.000	03:25
Air Blank		0.000	03:27
Subject Sample	# 2	0.170	03:27
Breath Volume		2.710	03:27
Air Blank		0.000	03:28
Control Test		0.078	03:28
Air Blank		0.000	03:29
Diagnostics Check		OK	03:29

Data Download Date/Time: 11/16/2017 14:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELSMORE,JUSTIN,R
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: MOGERLEY,NEIL
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 10/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 04:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:37

SUBJECT'S LAST NAME: FREDRICKSON

DATE/TIME OF TEST: 10/30/2017 05:04:11

Diagnostics Check	OK	04:58
Air Blank	0.000	04:58
Control Test	0.080	04:58
Air Blank	0.000	04:59
Subject Sample # 1	0.184	04:59
Breath Volume	2.710	04:59
Air Blank	0.000	05:00
Air Blank	0.000	05:02
Subject Sample # 2	0.189	05:02
Breath Volume	2.324	05:02
Air Blank	0.000	05:03
Control Test	0.079	05:03
Air Blank	0.000	05:04
Diagnostics Check	OK	05:04

Data Download Date/Time: 11/16/2017 14:24