Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REYES,GEORGE,
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: REYES,GEORGE
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 10/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 477627
DRY GAS EXPIRY DATE: 06/29/2019
VIOLATION CODE: DUI
ARREST TIME: 23:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:12

SUBJECT'S LAST NAME: FRANCISCO-GOMEZ

DATE/TIME OF TEST: 11/04/2017 23:46:58

Diagnostics Check		OK	23:38
Air Blank		0.000	23:39
Control Test		0.081	23:39
Air Blank		0.000	23:39
Subject Sample	# 1	0.182	23:42
Breath Volume		1.175	23:42
Air Blank		0.000	23:43
Air Blank		0.000	23:44
Subject Sample	# 2	0.202	23:45
Breath Volume		2.367	23:45
Air Blank		0.000	23:46
Control Test		0.080	23:46
Air Blank		0.000	23:46
Diagnostics Check		OK	23:46

Data Download Date/Time: 11/29/2017 19:00

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAYS,DARRYL,A
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: AGUIRRE,RICARDO
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 10/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 477627
DRY GAS EXPIRY DATE: 06/29/2019
VIOLATION CODE: DUI
ARREST TIME: 20:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:35

SUBJECT'S LAST NAME: AGUILAR

DATE/TIME OF TEST: 11/05/2017 21:11:26

Diagnostics Check		OK	21:00
Air Blank		0.000	21:01
Control Test		0.079	21:01
Air Blank		0.000	21:02
Subject Sample	# 1	0.288	21:04
Breath Volume		1.363	21:04
Air Blank		0.000	21:05
Air Blank		0.000	21:07
Subject Sample	# 2	0.291	21:09
Breath Volume		1.593	21:09
Air Blank		0.000	21:10
Control Test		0.077	21:10
Air Blank		0.000	21:11
Diagnostics Check		OK	21:11

Data Download Date/Time: 11/29/2017 19:00

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MURGUIDO,ALEJANDRO,
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: MURGUIDO,ALEJANDRO
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 10/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 477627
DRY GAS EXPIRY DATE: 06/29/2019
VIOLATION CODE: DUI
ARREST TIME: 20:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:35

SUBJECT'S LAST NAME: MANZANARES

DATE/TIME OF TEST: 11/10/2017 21:06:09

Diagnostics Check		OK	20:59
Air Blank		0.000	20:59
Control Test		0.082	21:00
Air Blank		0.000	21:00
Subject Sample	# 1	0.113	21:01
Breath Volume		3.460	21:01
Air Blank		0.000	21:02
Air Blank		0.000	21:03
Subject Sample	# 2	0.107	21:04
Breath Volume		3.359	21:04
Air Blank		0.000	21:05
Control Test		0.077	21:05
Air Blank		0.000	21:06
Diagnostics Check		OK	21:06

Data Download Date/Time: 11/29/2017 19:00

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,RAUL,
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: MAYS,DARRYL
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 10/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 477627
DRY GAS EXPIRY DATE: 06/29/2019
VIOLATION CODE: DUI
ARREST TIME: 23:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: ANTONIO

DATE/TIME OF TEST: 11/11/2017 00:45:29

Diagnostics Check		OK	00:37	
Air Blank		0.000	00:37	
Control Test		0.080	00:37	
Air Blank		0.000	00:38	
Subject Sample	# 1		00:41	No Sample Provided
Breath Volume		0.000	00:41	
Air Blank		0.000	00:41	
Air Blank		0.000	00:43	
Subject Sample	# 2		00:44	Subject Test Refused
Breath Volume		0.000	00:44	
Air Blank		0.000	00:44	
Control Test		0.081	00:44	
Air Blank		0.000	00:45	
Diagnostics Check		OK	00:45	

Data Download Date/Time: 11/29/2017 19:00

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,RAUL,
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: PEREZ,RAUL
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 10/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 477627
DRY GAS EXPIRY DATE: 06/29/2019
VIOLATION CODE: DUI
ARREST TIME: 03:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: AGUILAR MARTIN

DATE/TIME OF TEST: 11/11/2017 04:28:31

Diagnostics Check		OK	04:20
Air Blank		0.000	04:21
Control Test		0.081	04:21
Air Blank		0.000	04:22
Subject Sample	# 1	0.250	04:23
Breath Volume		2.570	04:23
Air Blank		0.000	04:24
Air Blank		0.000	04:25
Subject Sample	# 2	0.247	04:26
Breath Volume		2.480	04:26
Air Blank		0.000	04:27
Control Test		0.078	04:27
Air Blank		0.000	04:28
Diagnostics Check		OK	04:28

Data Download Date/Time: 11/29/2017 19:00

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,RAUL,
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: PEREZ,RAUL
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 10/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 477627
DRY GAS EXPIRY DATE: 06/29/2019
VIOLATION CODE: DUI
ARREST TIME: 03:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: BALAN PENA

DATE/TIME OF TEST: 11/24/2017 04:44:06

Diagnostics Check		OK	04:35	
Air Blank		0.000	04:36	
Control Test		0.080	04:36	
Air Blank		0.000	04:36	
Subject Sample	# 1		04:40	No Sample Provided
Breath Volume		0.000	04:40	
Air Blank		0.000	04:40	
Air Blank		0.000	04:42	
Subject Sample	# 2		04:42	Subject Test Refused
Breath Volume		0.000	04:42	
Air Blank		0.000	04:43	
Control Test		0.081	04:43	
Air Blank		0.000	04:44	
Diagnostics Check		OK	04:44	

Data Download Date/Time: 11/29/2017 19:00

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,RAUL,
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: PEREZ,RAUL
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 10/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 477627
DRY GAS EXPIRY DATE: 06/29/2019
VIOLATION CODE: DUI
ARREST TIME: 02:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: ABEJA

DATE/TIME OF TEST: 11/25/2017 03:29:30

Diagnostics Check		OK	03:22
Air Blank		0.000	03:23
Control Test		0.081	03:23
Air Blank		0.000	03:24
Subject Sample	# 1	0.087	03:24
Breath Volume		2.105	03:24
Air Blank		0.000	03:25
Air Blank		0.000	03:27
Subject Sample	# 2	0.084	03:27
Breath Volume		2.761	03:27
Air Blank		0.000	03:28
Control Test		0.081	03:28
Air Blank		0.000	03:29
Diagnostics Check		OK	03:29

Data Download Date/Time: 11/29/2017 19:00

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEPHENSON,JENNIFER,L
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: HURTADO,KENNETH
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 11/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 477627
DRY GAS EXPIRY DATE: 06/29/2019
VIOLATION CODE: DUI
ARREST TIME: 20:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:54

SUBJECT'S LAST NAME: PATINO VILLASENOR

DATE/TIME OF TEST: 11/29/2017 21:27:33

Diagnostics Check		OK	21:20
Air Blank		0.000	21:21
Control Test		0.081	21:21
Air Blank		0.000	21:22
Subject Sample	# 1	0.197	21:22
Breath Volume		2.324	21:22
Air Blank		0.000	21:23
Air Blank		0.000	21:25
Subject Sample	# 2	0.191	21:25
Breath Volume		2.902	21:25
Air Blank		0.000	21:26
Control Test		0.079	21:26
Air Blank		0.000	21:27
Diagnostics Check		OK	21:27

Data Download Date/Time: 12/27/2017 17:31

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006214

INSTRUMENT REGISTERED TO: OCALA POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOEDICKE,WILLIAM,R
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: JOEDICKE,WILLIAM
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 10/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:33

SUBJECT'S LAST NAME: STOVER

DATE/TIME OF TEST: 11/03/2017 22:59:40

Diagnostics Check		OK	22:53
Air Blank		0.000	22:53
Control Test		0.078	22:54
Air Blank		0.000	22:54
Subject Sample	# 1	0.140	22:54
Breath Volume		2.722	22:54
Air Blank		0.000	22:55
Air Blank		0.000	22:57
Subject Sample	# 2	0.136	22:58
Breath Volume		1.738	22:58
Air Blank		0.000	22:58
Control Test		0.078	22:59
Air Blank		0.000	22:59
Diagnostics Check		OK	22:59

Data Download Date/Time: 12/03/2017 20:54

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006214

INSTRUMENT REGISTERED TO: OCALA POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,GARY,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: MILLER,GARY
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 10/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:03

SUBJECT'S LAST NAME: SOUTHER

DATE/TIME OF TEST: 11/04/2017 02:33:00

Diagnostics Check		OK	02:26
Air Blank		0.000	02:27
Control Test		0.078	02:27
Air Blank		0.000	02:27
Subject Sample	# 1	0.143	02:28
Breath Volume		1.890	02:28
Air Blank		0.000	02:28
Air Blank		0.000	02:30
Subject Sample	# 2	0.146	02:31
Breath Volume		2.582	02:31
Air Blank		0.000	02:32
Control Test		0.077	02:32
Air Blank		0.000	02:32
Diagnostics Check		OK	02:33

Data Download Date/Time: 12/03/2017 20:54

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006214

INSTRUMENT REGISTERED TO: OCALA POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOEDICKE,WILLIAM,R
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: JOEDICKE,WILLIAM
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 10/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:44

SUBJECT'S LAST NAME: MILLS

DATE/TIME OF TEST: 11/05/2017 00:10:36

Diagnostics Check		OK	00:04
Air Blank		0.000	00:04
Control Test		0.078	00:05
Air Blank		0.000	00:05
Subject Sample	# 1	0.169	00:06
Breath Volume		2.277	00:06
Air Blank		0.000	00:06
Air Blank		0.000	00:08
Subject Sample	# 2	0.167	00:09
Breath Volume		2.863	00:09
Air Blank		0.000	00:09
Control Test		0.076	00:09
Air Blank		0.000	00:10
Diagnostics Check		OK	00:10

Data Download Date/Time: 12/03/2017 20:54

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006214

INSTRUMENT REGISTERED TO: OCALA POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOEDICKE,WILLIAM,R
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: JOEDICKE,WILLIAM
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 10/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:47

SUBJECT'S LAST NAME: COPENHAVER

DATE/TIME OF TEST: 11/11/2017 02:29:06

Diagnostics Check		OK	02:22
Air Blank		0.000	02:23
Control Test		0.078	02:23
Air Blank		0.000	02:24
Subject Sample	# 1	0.116	02:24
Breath Volume		3.000	02:24
Air Blank		0.000	02:25
Air Blank		0.000	02:27
Subject Sample	# 2	0.107	02:27
Breath Volume		1.671	02:27
Air Blank		0.000	02:28
Control Test		0.078	02:28
Air Blank		0.000	02:28
Diagnostics Check		OK	02:29

Data Download Date/Time: 12/03/2017 20:54

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006214

INSTRUMENT REGISTERED TO: OCALA POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOEDICKE,WILLIAM,R
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: JOEDICKE,WILLIAM
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 10/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:28

SUBJECT'S LAST NAME: ROVIRA

DATE/TIME OF TEST: 11/17/2017 23:50:02

Diagnostics Check	OK	23:49	
Air Blank		23:49	RFI Detect
Air Blank	0.000	23:50	

Data Download Date/Time: 12/03/2017 20:54

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006214

INSTRUMENT REGISTERED TO: OCALA POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOEDICKE,WILLIAM,R
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: JOEDICKE,WILLIAM
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 10/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:28

SUBJECT'S LAST NAME: ROVIRA

DATE/TIME OF TEST: 11/17/2017 23:57:30

Diagnostics Check		OK	23:52	
Air Blank		0.000	23:52	
Control Test		0.079	23:52	
Air Blank		0.000	23:53	
Subject Sample	# 1	0.115	23:54	
Breath Volume		1.902	23:54	
Air Blank		0.000	23:54	
Air Blank		0.000	23:56	
Subject Sample	# 2		23:56	RFI Detect
Breath Volume		0.550	23:56	
Air Blank			23:56	RFI Detect
Air Blank		0.000	23:57	

Data Download Date/Time: 12/03/2017 20:54

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006214

INSTRUMENT REGISTERED TO: OCALA POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOEDICKE,WILLIAM,R
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: JOEDICKE,WILLIAM
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 10/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:28

SUBJECT'S LAST NAME: ROVIRA

DATE/TIME OF TEST: 11/18/2017 00:06:48

Diagnostics Check		OK	00:00
Air Blank		0.000	00:00
Control Test		0.078	00:01
Air Blank		0.000	00:01
Subject Sample	# 1	0.109	00:02
Breath Volume		1.746	00:02
Air Blank		0.000	00:02
Air Blank		0.000	00:04
Subject Sample	# 2	0.118	00:05
Breath Volume		1.925	00:05
Air Blank		0.000	00:05
Control Test		0.077	00:06
Air Blank		0.000	00:06
Diagnostics Check		OK	00:06

Data Download Date/Time: 12/03/2017 20:54

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006214

INSTRUMENT REGISTERED TO: OCALA POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOEDICKE,WILLIAM,R
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: JOEDICKE,WILLIAM
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 10/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: HERNANDEZ VEGA

DATE/TIME OF TEST: 11/19/2017 02:53:36

Diagnostics Check		OK	02:47
Air Blank		0.000	02:47
Control Test		0.078	02:47
Air Blank		0.000	02:48
Subject Sample	# 1	0.174	02:49
Breath Volume		2.773	02:49
Air Blank		0.000	02:49
Air Blank		0.000	02:51
Subject Sample	# 2	0.176	02:52
Breath Volume		2.660	02:52
Air Blank		0.000	02:52
Control Test		0.077	02:52
Air Blank		0.000	02:53
Diagnostics Check		OK	02:53

Data Download Date/Time: 12/03/2017 20:55

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOORE,KIMBRLY,N
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: MOORE,KIMBERLY
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 11/16/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:13

SUBJECT'S LAST NAME: GROOMS

DATE/TIME OF TEST: 11/17/2017 21:10:21

Diagnostics Check		OK	21:03	
Air Blank		0.000	21:04	
Control Test		0.077	21:04	
Air Blank		0.000	21:05	
Subject Sample	# 1	0.036	21:05	
Breath Volume		1.855	21:05	
Air Blank		0.000	21:06	
Air Blank		0.000	21:08	
Subject Sample	# 2	0.031	21:08	
Breath Volume		1.765	21:08	
Air Blank		0.000	21:09	
Control Test		0.072	21:09	Control Outside Tolerance
Air Blank		0.000	21:10	

Data Download Date/Time: 01/02/2018 20:25

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOORE,KIMBERLY,N
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: MOORE,KIMBERLY
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 11/16/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:13

SUBJECT'S LAST NAME: GROOMS

DATE/TIME OF TEST: 11/17/2017 21:15:48

Diagnostics Check		OK	21:12	
Air Blank		0.000	21:13	
Control Test		0.077	21:13	
Air Blank		0.000	21:14	
Subject Sample	# 1		21:14	Subject Test Refused
Breath Volume		0.000	21:14	
Air Blank		0.000	21:14	
Control Test		0.078	21:15	
Air Blank		0.000	21:15	
Diagnostics Check		OK	21:15	

Data Download Date/Time: 01/02/2018 20:25

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOORE,KIMBERLY,N
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: MOORE,KIMBERLY
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 11/16/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:09

SUBJECT'S LAST NAME: MORROW

DATE/TIME OF TEST: 11/19/2017 02:37:00

Diagnostics Check		OK	02:30
Air Blank		0.000	02:31
Control Test		0.078	02:31
Air Blank		0.000	02:32
Subject Sample	# 1	0.176	02:32
Breath Volume		2.050	02:32
Air Blank		0.000	02:33
Air Blank		0.000	02:34
Subject Sample	# 2	0.176	02:35
Breath Volume		1.863	02:35
Air Blank		0.000	02:36
Control Test		0.076	02:36
Air Blank		0.000	02:36
Diagnostics Check		OK	02:37

Data Download Date/Time: 01/02/2018 20:25

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOORE,KIMBERLY,N
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: MOORE,KIMBERLY
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 11/16/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:06

SUBJECT'S LAST NAME: KIMBERLING

DATE/TIME OF TEST: 11/23/2017 01:47:51

Diagnostics Check		OK	01:41
Air Blank		0.000	01:41
Control Test		0.077	01:42
Air Blank		0.000	01:42
Subject Sample	# 1	0.186	01:43
Breath Volume		2.304	01:43
Air Blank		0.000	01:43
Air Blank		0.000	01:45
Subject Sample	# 2	0.173	01:46
Breath Volume		2.187	01:46
Air Blank		0.000	01:46
Control Test		0.077	01:47
Air Blank		0.000	01:47
Diagnostics Check		OK	01:47

Data Download Date/Time: 01/02/2018 20:25

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOORE,KIMBRLY,N
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: MOORE,KIMBERLY
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 11/16/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:53

SUBJECT'S LAST NAME: CLARK

DATE/TIME OF TEST: 11/25/2017 21:22:44

Diagnostics Check		OK	21:16
Air Blank		0.000	21:16
Control Test		0.078	21:16
Air Blank		0.000	21:17
Subject Sample	# 1	0.271	21:18
Breath Volume		1.914	21:18
Air Blank		0.000	21:18
Air Blank		0.000	21:20
Subject Sample	# 2	0.255	21:21
Breath Volume		1.761	21:21
Air Blank		0.000	21:21
Control Test		0.077	21:22
Air Blank		0.000	21:22
Diagnostics Check		OK	21:22

Data Download Date/Time: 01/02/2018 20:25

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006234

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMERON,DAVID,J
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: ANDERS,SHAWN
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 10/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 437738
DRY GAS EXPIRY DATE: 06/08/2019
VIOLATION CODE: DUI
ARREST TIME: 00:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: RODRIGUEZ VASQUEZ

DATE/TIME OF TEST: 11/12/2017 01:24:05

Diagnostics Check		OK	01:15
Air Blank		0.000	01:15
Control Test		0.081	01:16
Air Blank		0.000	01:16
Subject Sample	# 1	0.166	01:19
Breath Volume		2.378	01:19
Air Blank		0.000	01:20
Air Blank		0.000	01:21
Subject Sample	# 2	0.163	01:22
Breath Volume		2.125	01:22
Air Blank		0.000	01:23
Control Test		0.078	01:23
Air Blank		0.000	01:24
Diagnostics Check		OK	01:24

Data Download Date/Time: 11/19/2017 23:54

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006234

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMERON,DAVID,J
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: CAMERON,DAVID
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 11/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 437738
DRY GAS EXPIRY DATE: 06/08/2019
VIOLATION CODE: DUI
ARREST TIME: 02:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:32

SUBJECT'S LAST NAME: WHEELER

DATE/TIME OF TEST: 11/26/2017 04:10:24

Diagnostics Check		OK	04:03
Air Blank		0.000	04:04
Control Test		0.081	04:04
Air Blank		0.000	04:05
Subject Sample	# 1	0.000	04:05
Breath Volume		1.542	04:05
Air Blank		0.000	04:06
Air Blank		0.000	04:08
Subject Sample	# 2	0.000	04:09
Breath Volume		1.218	04:09
Air Blank		0.000	04:09
Control Test		0.080	04:09
Air Blank		0.000	04:10
Diagnostics Check		OK	04:10

Data Download Date/Time: 12/26/2017 19:37

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006234

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLEURANT,RODSON,
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: FLEURANT,RODSON
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 11/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 437738
DRY GAS EXPIRY DATE: 06/08/2019
VIOLATION CODE: DUI
ARREST TIME: 16:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:58

SUBJECT'S LAST NAME: SCHALL

DATE/TIME OF TEST: 11/30/2017 17:44:36

Diagnostics Check		OK	17:38
Air Blank		0.000	17:38
Control Test		0.081	17:39
Air Blank		0.000	17:39
Subject Sample	# 1	0.050	17:40
Breath Volume		1.980	17:40
Air Blank		0.000	17:40
Air Blank		0.000	17:42
Subject Sample	# 2	0.052	17:43
Breath Volume		2.148	17:43
Air Blank		0.000	17:43
Control Test		0.081	17:43
Air Blank		0.000	17:44
Diagnostics Check		OK	17:44

Data Download Date/Time: 12/26/2017 19:37

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROWE,BRANDON,D
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: ROWE,BRANDON
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 10/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 00:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: MARTINEZ HERNANDEZ

DATE/TIME OF TEST: 11/08/2017 01:19:29

Diagnostics Check		OK	01:11
Air Blank		0.000	01:11
Control Test		0.080	01:12
Air Blank		0.000	01:12
Subject Sample	# 1	0.154	01:14
Breath Volume		1.617	01:14
Air Blank		0.000	01:15
Air Blank		0.000	01:16
Subject Sample	# 2	0.143	01:17
Breath Volume		2.058	01:17
Air Blank		0.000	01:18
Control Test		0.079	01:18
Air Blank		0.000	01:19
Diagnostics Check		OK	01:19

Data Download Date/Time: 11/27/2017 10:00

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROWE,BRANDON,D
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: MERRILL,JACOB
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 10/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 22:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:33

SUBJECT'S LAST NAME: SEROV

DATE/TIME OF TEST: 11/12/2017 23:20:23

Diagnostics Check		OK	23:07
Air Blank		0.000	23:08
Control Test		0.081	23:08
Air Blank		0.000	23:09
Subject Sample	# 1	0.202	23:11
Breath Volume		2.292	23:11
Air Blank		0.000	23:12
Air Blank		0.000	23:13
Subject Sample	# 2	0.181	23:15
Breath Volume		1.625	23:15
Air Blank		0.000	23:16
Air Blank		0.000	23:17
Subject Sample	# 3	0.189	23:18
Breath Volume		3.066	23:18
Air Blank		0.000	23:19
Control Test		0.079	23:19
Air Blank		0.000	23:20
Diagnostics Check		OK	23:20

Data Download Date/Time: 11/27/2017 10:00

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROWE,BRANDON,D
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: ROWE,BRANDON
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 10/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 22:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: VELASQUEZ LOPEZ

DATE/TIME OF TEST: 11/12/2017 23:29:03

Diagnostics Check		OK	23:22
Air Blank		0.000	23:23
Control Test		0.079	23:23
Air Blank		0.000	23:23
Subject Sample	# 1	0.130	23:24
Breath Volume		2.914	23:24
Air Blank		0.000	23:25
Air Blank		0.000	23:26
Subject Sample	# 2	0.122	23:27
Breath Volume		2.085	23:27
Air Blank		0.000	23:28
Control Test		0.079	23:28
Air Blank		0.000	23:28
Diagnostics Check		OK	23:29

Data Download Date/Time: 11/27/2017 10:00

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROWE,BRANDON,D
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: ROWE,BRANDON
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 10/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 01:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:36

SUBJECT'S LAST NAME: MAUGHERMAN

DATE/TIME OF TEST: 11/16/2017 02:02:40

Diagnostics Check		OK	01:56
Air Blank		0.000	01:56
Control Test		0.080	01:57
Air Blank		0.000	01:57
Subject Sample	# 1	0.188	01:57
Breath Volume		2.332	01:57
Air Blank		0.000	01:58
Air Blank		0.000	02:00
Subject Sample	# 2	0.188	02:00
Breath Volume		2.734	02:00
Air Blank		0.000	02:01
Control Test		0.080	02:02
Air Blank		0.000	02:02
Diagnostics Check		OK	02:02

Data Download Date/Time: 11/27/2017 10:00

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROWE,BRANDON,D
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: ROWE,BRANDON
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 10/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 01:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:13

SUBJECT'S LAST NAME: GARAY BONILLA

DATE/TIME OF TEST: 11/21/2017 01:41:06

Diagnostics Check		OK	01:34
Air Blank		0.000	01:35
Control Test		0.082	01:35
Air Blank		0.000	01:36
Subject Sample	# 1	0.152	01:36
Breath Volume		2.546	01:36
Air Blank		0.000	01:37
Air Blank		0.000	01:39
Subject Sample	# 2	0.146	01:39
Breath Volume		2.656	01:39
Air Blank		0.000	01:40
Control Test		0.079	01:40
Air Blank		0.000	01:41
Diagnostics Check		OK	01:41

Data Download Date/Time: 11/27/2017 10:00

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROWE,BRANDON,D
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: RIVERA,LUIS
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 10/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 22:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: TRISTAN SANCHEZ

DATE/TIME OF TEST: 11/26/2017 23:42:33

Diagnostics Check		OK	23:35
Air Blank		0.000	23:35
Control Test		0.082	23:36
Air Blank		0.000	23:36
Subject Sample	# 1	0.137	23:37
Breath Volume		1.355	23:37
Air Blank		0.000	23:38
Air Blank		0.000	23:40
Subject Sample	# 2	0.138	23:40
Breath Volume		1.234	23:40
Air Blank		0.000	23:41
Control Test		0.081	23:41
Air Blank		0.000	23:42
Diagnostics Check		OK	23:42

Data Download Date/Time: 11/27/2017 10:01

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROWE,BRANDON,D
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: ROWE,BRANDON
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 11/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:24

SUBJECT'S LAST NAME: SILER

DATE/TIME OF TEST: 11/30/2017 01:14:29

Diagnostics Check		OK	01:08
Air Blank		0.000	01:08
Control Test		0.081	01:09
Air Blank		0.000	01:09
Subject Sample	# 1	0.122	01:10
Breath Volume		3.132	01:10
Air Blank		0.000	01:10
Air Blank		0.000	01:12
Subject Sample	# 2	0.127	01:12
Breath Volume		2.574	01:12
Air Blank		0.000	01:13
Control Test		0.081	01:13
Air Blank		0.000	01:14
Diagnostics Check		OK	01:14

Data Download Date/Time: 12/18/2017 11:07

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROWE,BRANDON,D
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: ROWE,BRANDON
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 11/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 18:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:36

SUBJECT'S LAST NAME: KYTE

DATE/TIME OF TEST: 11/30/2017 19:03:44

Diagnostics Check		OK	18:57
Air Blank		0.000	18:58
Control Test		0.080	18:58
Air Blank		0.000	18:58
Subject Sample	# 1	0.000	18:59
Breath Volume		3.457	18:59
Air Blank		0.000	18:59
Air Blank		0.000	19:01
Subject Sample	# 2	0.000	19:02
Breath Volume		3.035	19:02
Air Blank		0.000	19:02
Control Test		0.080	19:03
Air Blank		0.000	19:03
Diagnostics Check		OK	19:03

Data Download Date/Time: 12/18/2017 11:07

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,DAKOTA,W
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: WHITE,M
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 10/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: PAULK

DATE/TIME OF TEST: 11/01/2017 01:12:29

Diagnostics Check		OK	01:05
Air Blank		0.000	01:06
Control Test		0.080	01:06
Air Blank		0.000	01:06
Subject Sample	# 1	0.114	01:07
Breath Volume		2.652	01:07
Air Blank		0.000	01:08
Air Blank		0.000	01:10
Subject Sample	# 2	0.108	01:10
Breath Volume		2.816	01:10
Air Blank		0.000	01:11
Control Test		0.080	01:11
Air Blank		0.000	01:12
Diagnostics Check		OK	01:12

Data Download Date/Time: 11/21/2017 11:29

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,DAKOTA,W
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: LANIER,H
ARRESTING OFFICER AGENCY: Orange Park PD
LAST AGENCY INSPECTION: 10/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 18:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:55

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 11/03/2017 19:27:04

Diagnostics Check		OK	19:15	
Air Blank		0.000	19:16	
Control Test		0.081	19:16	
Air Blank		0.000	19:16	
Subject Sample	# 1		19:20	No Sample Provided
Breath Volume		0.000	19:20	
Air Blank		0.000	19:20	
Air Blank		0.000	19:22	
Subject Sample	# 2		19:25	No Sample Provided
Breath Volume		0.000	19:25	
Air Blank		0.000	19:26	
Control Test		0.080	19:26	
Air Blank		0.000	19:26	
Diagnostics Check		OK	19:27	

Data Download Date/Time: 11/21/2017 11:29

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,DAKOTA,W
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: COLLINS,K
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 10/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:58

SUBJECT'S LAST NAME: MONDAY

DATE/TIME OF TEST: 11/03/2017 22:32:08

Diagnostics Check		OK	22:18	
Air Blank		0.000	22:19	
Control Test		0.080	22:19	
Air Blank		0.000	22:19	
Subject Sample	# 1	0.082	22:23	Volume Not Met
Breath Volume		0.511	22:23	
Air Blank		0.000	22:23	
Air Blank		0.000	22:25	
Subject Sample	# 2	0.103	22:26	
Breath Volume		2.148	22:26	
Air Blank		0.000	22:27	
Air Blank		0.000	22:28	
Subject Sample	# 3	0.100	22:30	
Breath Volume		1.617	22:30	
Air Blank		0.000	22:31	
Control Test		0.079	22:31	
Air Blank		0.000	22:32	
Diagnostics Check		OK	22:32	

Data Download Date/Time: 11/21/2017 11:29

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,DAKOTA,W
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: WARREN,C
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 10/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: CRIBBS

DATE/TIME OF TEST: 11/04/2017 03:35:09

Diagnostics Check		OK	03:28
Air Blank		0.000	03:29
Control Test		0.081	03:29
Air Blank		0.000	03:30
Subject Sample	# 1	0.208	03:30
Breath Volume		1.476	03:30
Air Blank		0.000	03:31
Air Blank		0.000	03:32
Subject Sample	# 2	0.203	03:33
Breath Volume		2.375	03:33
Air Blank		0.000	03:34
Control Test		0.079	03:34
Air Blank		0.000	03:35
Diagnostics Check		OK	03:35

Data Download Date/Time: 11/21/2017 11:29

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,DAKOTA,W
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: BAHNSEN,A
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 10/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:01

SUBJECT'S LAST NAME: WARD

DATE/TIME OF TEST: 11/05/2017 01:33:10

Diagnostics Check		OK	01:21	
Air Blank		0.000	01:22	
Control Test		0.080	01:22	
Air Blank		0.000	01:23	
Subject Sample	# 1		01:26	No Sample Provided
Breath Volume		0.000	01:26	
Air Blank		0.000	01:26	
Air Blank		0.000	01:28	
Subject Sample	# 2		01:31	No Sample Provided
Breath Volume		0.000	01:31	
Air Blank		0.000	01:32	
Control Test		0.081	01:32	
Air Blank		0.000	01:33	
Diagnostics Check		OK	01:33	

Data Download Date/Time: 11/21/2017 11:29

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STACY,A,L
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: STACY,A
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 10/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: ADMIN
ARREST TIME: 08:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:57

SUBJECT'S LAST NAME: BURCHETT

DATE/TIME OF TEST: 11/10/2017 09:23:10

Diagnostics Check		OK	09:17
Air Blank		0.000	09:17
Control Test		0.081	09:17
Air Blank		0.000	09:18
Subject Sample	# 1	0.000	09:18
Breath Volume		2.265	09:18
Air Blank		0.000	09:19
Air Blank		0.000	09:21
Subject Sample	# 2	0.000	09:21
Breath Volume		2.250	09:21
Air Blank		0.000	09:22
Control Test		0.081	09:22
Air Blank		0.000	09:23
Diagnostics Check		OK	09:23

Data Download Date/Time: 11/21/2017 11:30

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,DAKOTA,W
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: YOUNGBLOOD,C
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 10/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:03

SUBJECT'S LAST NAME: PEREZ PEREZ

DATE/TIME OF TEST: 11/14/2017 03:33:13

Diagnostics Check		OK	03:26
Air Blank		0.000	03:26
Control Test		0.080	03:27
Air Blank		0.000	03:27
Subject Sample	# 1	0.186	03:28
Breath Volume		2.910	03:28
Air Blank		0.000	03:29
Air Blank		0.000	03:31
Subject Sample	# 2	0.194	03:31
Breath Volume		1.804	03:31
Air Blank		0.000	03:32
Control Test		0.079	03:32
Air Blank		0.000	03:33
Diagnostics Check		OK	03:33

Data Download Date/Time: 11/21/2017 11:30

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINSON,C,A
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: TENNANT,R
ARRESTING OFFICER AGENCY: Orange Park PD
LAST AGENCY INSPECTION: 10/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: JOHNSON-DUKE

DATE/TIME OF TEST: 11/16/2017 01:43:05

Diagnostics Check	OK	01:41	
Air Blank	0.000	01:42	
Control Test	0.080	01:42	
Air Blank		01:42	RFI Detect
Air Blank	0.000	01:43	

Data Download Date/Time: 11/21/2017 11:30

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINSON,C,A
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: TENNANT,R
ARRESTING OFFICER AGENCY: Orange Park PD
LAST AGENCY INSPECTION: 10/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: JOHNSON-DUKE

DATE/TIME OF TEST: 11/16/2017 01:47:32

Diagnostics Check		OK	01:44	
Air Blank		0.000	01:45	
Control Test		0.081	01:45	
Air Blank		0.000	01:45	
Subject Sample	# 1		01:46	Subject Test Refused
Breath Volume		0.000	01:46	
Air Blank		0.000	01:46	
Control Test		0.081	01:46	
Air Blank		0.000	01:47	
Diagnostics Check		OK	01:47	

Data Download Date/Time: 11/21/2017 11:30

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAIN,JEANETTE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: YOCHUM,CRAIG
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 10/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: BARRIO CAX

DATE/TIME OF TEST: 11/01/2017 01:22:02

Diagnostics Check		OK	01:15
Air Blank		0.000	01:16
Control Test		0.080	01:16
Air Blank		0.000	01:17
Subject Sample	# 1	0.166	01:17
Breath Volume		3.437	01:17
Air Blank		0.000	01:18
Air Blank		0.000	01:20
Subject Sample	# 2	0.163	01:20
Breath Volume		3.386	01:20
Air Blank		0.000	01:21
Control Test		0.080	01:21
Air Blank		0.000	01:21
Diagnostics Check		OK	01:22

Data Download Date/Time: 11/17/2017 15:27

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARI,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: PHILIPPE,BEDJY
ARRESTING OFFICER AGENCY: Delray Beach PD
LAST AGENCY INSPECTION: 10/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: SILLER

DATE/TIME OF TEST: 11/02/2017 03:27:33

Diagnostics Check		OK	03:21
Air Blank		0.000	03:21
Control Test		0.081	03:22
Air Blank		0.000	03:22
Subject Sample	# 1	0.173	03:22
Breath Volume		2.707	03:22
Air Blank		0.000	03:23
Air Blank		0.000	03:25
Subject Sample	# 2	0.173	03:26
Breath Volume		2.937	03:26
Air Blank		0.000	03:26
Control Test		0.080	03:26
Air Blank		0.000	03:27
Diagnostics Check		OK	03:27

Data Download Date/Time: 11/17/2017 15:27

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAIN,JEANETTE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MUNOZ,NICOLE
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 10/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:45

SUBJECT'S LAST NAME: RAMIREZ

DATE/TIME OF TEST: 11/02/2017 21:17:55

Diagnostics Check		OK	21:11
Air Blank		0.000	21:12
Control Test		0.080	21:12
Air Blank		0.000	21:12
Subject Sample	# 1	0.178	21:13
Breath Volume		1.765	21:13
Air Blank		0.000	21:14
Air Blank		0.000	21:15
Subject Sample	# 2	0.174	21:16
Breath Volume		1.359	21:16
Air Blank		0.000	21:17
Control Test		0.080	21:17
Air Blank		0.000	21:17
Diagnostics Check		OK	21:17

Data Download Date/Time: 11/17/2017 15:27

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARI,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITE,EDWARD
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 10/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:28

SUBJECT'S LAST NAME: HARRELL

DATE/TIME OF TEST: 11/04/2017 03:11:57

Diagnostics Check		OK	02:57	
Air Blank		0.000	02:57	
Control Test		0.080	02:58	
Air Blank		0.000	02:58	
Subject Sample	# 1	0.248	03:01	
Breath Volume		2.085	03:01	
Air Blank		0.000	03:01	
Air Blank		0.000	03:03	
Subject Sample	# 2	0.209	03:06	Volume Not Met
Breath Volume		0.707	03:06	
Air Blank		0.000	03:07	
Air Blank		0.000	03:09	
Subject Sample	# 3	0.236	03:10	
Breath Volume		1.183	03:10	
Air Blank		0.000	03:11	
Control Test		0.076	03:11	
Air Blank		0.000	03:11	
Diagnostics Check		OK	03:11	

Data Download Date/Time: 11/17/2017 15:27

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MACKEY,JAMES
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 10/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 07:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:43

SUBJECT'S LAST NAME: STERLING

DATE/TIME OF TEST: 11/04/2017 08:22:54

Diagnostics Check		OK	08:11	
Air Blank		0.000	08:11	
Control Test		0.080	08:11	
Air Blank		0.000	08:12	
Subject Sample	# 1	0.241	08:15	Volume Not Met
Breath Volume		0.117	08:15	
Air Blank		0.000	08:16	
Air Blank		0.000	08:18	
Subject Sample	# 2	0.218	08:21	Volume Not Met
Breath Volume		0.660	08:21	
Air Blank		0.000	08:21	
Control Test		0.078	08:22	
Air Blank		0.000	08:22	
Diagnostics Check		OK	08:22	

Data Download Date/Time: 11/17/2017 15:27

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MACKEY,JAMES
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 10/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 07:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:43

SUBJECT'S LAST NAME: STERLING

DATE/TIME OF TEST: 11/04/2017 08:33:24

Diagnostics Check		OK	08:25
Air Blank		0.000	08:26
Control Test		0.078	08:26
Air Blank		0.000	08:27
Subject Sample	# 1	0.255	08:28
Breath Volume		2.046	08:28
Air Blank		0.000	08:29
Air Blank		0.000	08:31
Subject Sample	# 2	0.255	08:31
Breath Volume		1.269	08:31
Air Blank		0.000	08:32
Control Test		0.079	08:32
Air Blank		0.000	08:33
Diagnostics Check		OK	08:33

Data Download Date/Time: 11/17/2017 15:27

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: TRENTMAN,THOMAS
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 10/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 08:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:48

SUBJECT'S LAST NAME: MARTIN RAMIREZ

DATE/TIME OF TEST: 11/04/2017 09:24:05

Diagnostics Check		OK	09:15
Air Blank		0.000	09:15
Control Test		0.080	09:16
Air Blank		0.000	09:16
Subject Sample	# 1	0.192	09:18
Breath Volume		1.351	09:18
Air Blank		0.000	09:19
Air Blank		0.000	09:21
Subject Sample	# 2	0.177	09:22
Breath Volume		1.449	09:22
Air Blank		0.000	09:23
Control Test		0.078	09:23
Air Blank		0.000	09:23
Diagnostics Check		OK	09:24

Data Download Date/Time: 11/17/2017 15:27

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CARMENATE,JOSHUA
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 10/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 10:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:15

SUBJECT'S LAST NAME: BILLINGS

DATE/TIME OF TEST: 11/04/2017 11:48:20

Diagnostics Check		OK	11:40
Air Blank		0.000	11:41
Control Test		0.080	11:41
Air Blank		0.000	11:41
Subject Sample	# 1	0.198	11:43
Breath Volume		1.468	11:43
Air Blank		0.000	11:43
Air Blank		0.000	11:45
Subject Sample	# 2	0.202	11:46
Breath Volume		1.566	11:46
Air Blank		0.000	11:47
Control Test		0.079	11:47
Air Blank		0.000	11:48
Diagnostics Check		OK	11:48

Data Download Date/Time: 11/17/2017 15:27

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SALM,PHILIP
ARRESTING OFFICER AGENCY: Ocean Ridge PD
LAST AGENCY INSPECTION: 10/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 15:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:40

SUBJECT'S LAST NAME: MCCARTHY

DATE/TIME OF TEST: 11/04/2017 16:46:30

Diagnostics Check		OK	16:39
Air Blank		0.000	16:40
Control Test		0.080	16:40
Air Blank		0.000	16:41
Subject Sample	# 1	0.065	16:41
Breath Volume		2.937	16:41
Air Blank		0.000	16:42
Air Blank		0.000	16:44
Subject Sample	# 2	0.064	16:44
Breath Volume		2.878	16:44
Air Blank		0.000	16:45
Control Test		0.079	16:45
Air Blank		0.000	16:46
Diagnostics Check		OK	16:46

Data Download Date/Time: 11/17/2017 15:27

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: HANTON,MELINDA
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 10/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:28

SUBJECT'S LAST NAME: SPENCER

DATE/TIME OF TEST: 11/04/2017 22:00:21

Diagnostics Check		OK	21:53
Air Blank		0.000	21:54
Control Test		0.081	21:54
Air Blank		0.000	21:55
Subject Sample	# 1	0.132	21:55
Breath Volume		1.933	21:55
Air Blank		0.000	21:56
Air Blank		0.000	21:58
Subject Sample	# 2	0.128	21:58
Breath Volume		1.156	21:58
Air Blank		0.000	21:59
Control Test		0.080	21:59
Air Blank		0.000	22:00
Diagnostics Check		OK	22:00

Data Download Date/Time: 11/17/2017 15:27

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SCARTOZZI,PATRICK
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 10/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:43

SUBJECT'S LAST NAME: IXTABALAN LOPEZ

DATE/TIME OF TEST: 11/05/2017 00:20:30

Diagnostics Check		OK	00:11
Air Blank		0.000	00:12
Control Test		0.080	00:12
Air Blank		0.000	00:13
Subject Sample	# 1	0.174	00:15
Breath Volume		2.378	00:15
Air Blank		0.000	00:16
Air Blank		0.000	00:17
Subject Sample	# 2	0.169	00:18
Breath Volume		2.957	00:18
Air Blank		0.000	00:19
Control Test		0.078	00:19
Air Blank		0.000	00:20
Diagnostics Check		OK	00:20

Data Download Date/Time: 11/17/2017 15:27

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LEHENY,JACOB
ARRESTING OFFICER AGENCY: Palm Springs PD
LAST AGENCY INSPECTION: 10/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: GUOX VICENTE

DATE/TIME OF TEST: 11/05/2017 01:49:11

Diagnostics Check		OK	01:39
Air Blank		0.000	01:39
Control Test		0.080	01:39
Air Blank		0.000	01:40
Subject Sample	# 1	0.206	01:42
Breath Volume		1.156	01:42
Air Blank		0.000	01:43
Air Blank		0.000	01:45
Subject Sample	# 2	0.192	01:47
Breath Volume		1.281	01:47
Air Blank		0.000	01:48
Control Test		0.079	01:48
Air Blank		0.000	01:49
Diagnostics Check		OK	01:49

Data Download Date/Time: 11/17/2017 15:27

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SOLOMON,GEORGE
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 10/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:37

SUBJECT'S LAST NAME: VELASCO SANTIAGO

DATE/TIME OF TEST: 11/05/2017 02:32:31

Diagnostics Check		OK	02:25
Air Blank		0.000	02:26
Control Test		0.080	02:26
Air Blank		0.000	02:27
Subject Sample	# 1	0.148	02:27
Breath Volume		1.675	02:27
Air Blank		0.000	02:28
Air Blank		0.000	02:30
Subject Sample	# 2	0.139	02:30
Breath Volume		1.343	02:30
Air Blank		0.000	02:31
Control Test		0.079	02:31
Air Blank		0.000	02:32
Diagnostics Check		OK	02:32

Data Download Date/Time: 11/17/2017 15:27

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: TAVARES,LUCAS
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 10/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:18

SUBJECT'S LAST NAME: TORRENTES ZAMORA

DATE/TIME OF TEST: 11/05/2017 03:56:00

Diagnostics Check		OK	03:49
Air Blank		0.000	03:49
Control Test		0.080	03:49
Air Blank		0.000	03:50
Subject Sample	# 1	0.129	03:51
Breath Volume		2.367	03:51
Air Blank		0.000	03:52
Air Blank		0.000	03:53
Subject Sample	# 2	0.131	03:54
Breath Volume		2.792	03:54
Air Blank		0.000	03:55
Control Test		0.079	03:55
Air Blank		0.000	03:55
Diagnostics Check		OK	03:56

Data Download Date/Time: 11/17/2017 15:27

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAIN,JEANETTE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: PHILIPPE,B
ARRESTING OFFICER AGENCY: Delray Beach PD
LAST AGENCY INSPECTION: 10/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: COTTRELL

DATE/TIME OF TEST: 11/07/2017 22:39:53

Diagnostics Check		OK	22:37	
Air Blank		0.000	22:37	
Control Test		0.081	22:37	
Air Blank		0.000	22:38	
Subject Sample	# 1		22:38	Subject Test Refused
Breath Volume		0.000	22:38	
Air Blank		0.000	22:38	
Control Test		0.080	22:39	
Air Blank		0.000	22:39	
Diagnostics Check		OK	22:39	

Data Download Date/Time: 11/17/2017 15:27

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: QUESADA,YOSVANI
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 10/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: AYLWARD

DATE/TIME OF TEST: 11/07/2017 23:59:49

Diagnostics Check		OK	23:52
Air Blank		0.000	23:53
Control Test		0.080	23:53
Air Blank		0.000	23:53
Subject Sample	# 1	0.311	23:55
Breath Volume		1.816	23:55
Air Blank		0.000	23:55
Air Blank		0.000	23:57
Subject Sample	# 2	0.304	23:58
Breath Volume		1.632	23:58
Air Blank		0.000	23:58
Control Test		0.080	23:59
Air Blank		0.000	23:59
Diagnostics Check		OK	23:59

Data Download Date/Time: 11/17/2017 15:27

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAIN,JEANETTE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WALTON,THOMAS
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 10/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:53

SUBJECT'S LAST NAME: BURDESHAW

DATE/TIME OF TEST: 11/08/2017 21:22:31

Diagnostics Check		OK	21:16
Air Blank		0.000	21:16
Control Test		0.080	21:17
Air Blank		0.000	21:17
Subject Sample	# 1	0.000	21:18
Breath Volume		1.640	21:18
Air Blank		0.000	21:18
Air Blank		0.000	21:20
Subject Sample	# 2	0.000	21:21
Breath Volume		1.652	21:21
Air Blank		0.000	21:21
Control Test		0.080	21:21
Air Blank		0.000	21:22
Diagnostics Check		OK	21:22

Data Download Date/Time: 11/17/2017 15:27

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARI,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SALVEMINI,MICHAEL
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 10/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:15

SUBJECT'S LAST NAME: HOCHMAN

DATE/TIME OF TEST: 11/08/2017 22:46:34

Diagnostics Check		OK	22:40
Air Blank		0.000	22:40
Control Test		0.080	22:40
Air Blank		0.000	22:41
Subject Sample	# 1	0.190	22:42
Breath Volume		1.222	22:42
Air Blank		0.000	22:42
Air Blank		0.000	22:44
Subject Sample	# 2	0.188	22:45
Breath Volume		2.445	22:45
Air Blank		0.000	22:45
Control Test		0.080	22:45
Air Blank		0.000	22:46
Diagnostics Check		OK	22:46

Data Download Date/Time: 11/17/2017 15:27

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAIN,JEANETTE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITE,EDWARD KEITH
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 10/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: HUMPHREY

DATE/TIME OF TEST: 11/09/2017 01:32:09

Diagnostics Check		OK	01:24
Air Blank		0.000	01:24
Control Test		0.080	01:24
Air Blank		0.000	01:25
Subject Sample	# 1	0.272	01:27
Breath Volume		2.351	01:27
Air Blank		0.000	01:28
Air Blank		0.000	01:30
Subject Sample	# 2	0.264	01:30
Breath Volume		2.238	01:30
Air Blank		0.000	01:31
Control Test		0.079	01:31
Air Blank		0.000	01:32
Diagnostics Check		OK	01:32

Data Download Date/Time: 11/17/2017 15:27

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARI,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WALTON,THOMAS
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 10/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: NELSON

DATE/TIME OF TEST: 11/09/2017 03:32:42

Diagnostics Check		OK	03:24
Air Blank		0.000	03:25
Control Test		0.080	03:25
Air Blank		0.000	03:26
Subject Sample	# 1	0.149	03:27
Breath Volume		1.988	03:27
Air Blank		0.000	03:28
Air Blank		0.000	03:30
Subject Sample	# 2	0.146	03:31
Breath Volume		1.996	03:31
Air Blank		0.000	03:31
Control Test		0.077	03:32
Air Blank		0.000	03:32
Diagnostics Check		OK	03:32

Data Download Date/Time: 11/17/2017 15:27

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAIN,JEANETTE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: HANTON,MELINDA
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 10/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 14:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:35

SUBJECT'S LAST NAME: CRANE

DATE/TIME OF TEST: 11/09/2017 16:05:31

Diagnostics Check		OK	15:59
Air Blank		0.000	15:59
Control Test		0.081	15:59
Air Blank		0.000	16:00
Subject Sample	# 1	0.271	16:00
Breath Volume		3.562	16:00
Air Blank		0.000	16:01
Air Blank		0.000	16:03
Subject Sample	# 2	0.269	16:03
Breath Volume		4.777	16:03
Air Blank		0.000	16:04
Control Test		0.079	16:04
Air Blank		0.000	16:05
Diagnostics Check		OK	16:05

Data Download Date/Time: 11/17/2017 15:27

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARI,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: TODD,ZION
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 10/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: MULLETT

DATE/TIME OF TEST: 11/11/2017 04:16:33

Diagnostics Check		OK	04:08
Air Blank		0.000	04:09
Control Test		0.080	04:09
Air Blank		0.000	04:09
Subject Sample	# 1	0.177	04:11
Breath Volume		1.132	04:11
Air Blank		0.000	04:11
Air Blank		0.000	04:13
Subject Sample	# 2	0.192	04:14
Breath Volume		1.722	04:14
Air Blank		0.000	04:15
Control Test		0.080	04:15
Air Blank		0.000	04:16
Diagnostics Check		OK	04:16

Data Download Date/Time: 11/17/2017 15:27

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARI,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ROTHENBURG,JONATHAN
ARRESTING OFFICER AGENCY: Palm Beach PD
LAST AGENCY INSPECTION: 10/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:59

SUBJECT'S LAST NAME: RUBIO

DATE/TIME OF TEST: 11/11/2017 04:43:37

Diagnostics Check		OK	04:37
Air Blank		0.000	04:37
Control Test		0.080	04:38
Air Blank		0.000	04:38
Subject Sample	# 1	0.115	04:39
Breath Volume		1.648	04:39
Air Blank		0.000	04:39
Air Blank		0.000	04:41
Subject Sample	# 2	0.110	04:42
Breath Volume		1.968	04:42
Air Blank		0.000	04:42
Control Test		0.079	04:43
Air Blank		0.000	04:43
Diagnostics Check		OK	04:43

Data Download Date/Time: 11/17/2017 15:27

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARI,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ROTHENBURG,JONATHAN
ARRESTING OFFICER AGENCY: Palm Beach PD
LAST AGENCY INSPECTION: 10/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: RODRIGUEZ SERRANO

DATE/TIME OF TEST: 11/12/2017 00:33:25

Diagnostics Check		OK	00:27
Air Blank		0.000	00:27
Control Test		0.081	00:27
Air Blank		0.000	00:28
Subject Sample	# 1	0.209	00:28
Breath Volume		2.863	00:28
Air Blank		0.000	00:29
Air Blank		0.000	00:31
Subject Sample	# 2	0.216	00:31
Breath Volume		2.957	00:31
Air Blank		0.000	00:32
Control Test		0.079	00:32
Air Blank		0.000	00:33
Diagnostics Check		OK	00:33

Data Download Date/Time: 11/17/2017 15:27

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARI,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LEVEY,SCOTT
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 10/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:38

SUBJECT'S LAST NAME: ALVARADO

DATE/TIME OF TEST: 11/12/2017 01:19:24

Diagnostics Check		OK	01:13
Air Blank		0.000	01:13
Control Test		0.079	01:14
Air Blank		0.000	01:14
Subject Sample	# 1	0.214	01:15
Breath Volume		2.476	01:15
Air Blank		0.000	01:15
Air Blank		0.000	01:17
Subject Sample	# 2	0.206	01:17
Breath Volume		2.125	01:17
Air Blank		0.000	01:18
Control Test		0.079	01:18
Air Blank		0.000	01:19
Diagnostics Check		OK	01:19

Data Download Date/Time: 11/17/2017 15:27

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARI,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SCHAEFER,JOHN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 10/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: NOVOA

DATE/TIME OF TEST: 11/12/2017 03:15:06

Diagnostics Check		OK	03:03	
Air Blank		0.000	03:04	
Control Test		0.080	03:04	
Air Blank		0.000	03:04	
Subject Sample	# 1		03:08	No Sample Provided
Breath Volume		0.000	03:08	
Air Blank		0.000	03:08	
Air Blank		0.000	03:10	
Subject Sample	# 2		03:13	No Sample Provided
Breath Volume		0.000	03:13	
Air Blank		0.000	03:14	
Control Test		0.080	03:14	
Air Blank		0.000	03:15	
Diagnostics Check		OK	03:15	

Data Download Date/Time: 11/17/2017 15:27

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARI,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: TAVARES,LUCAS
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 10/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: ALVARADO

DATE/TIME OF TEST: 11/12/2017 03:43:19

Diagnostics Check		OK	03:37
Air Blank		0.000	03:37
Control Test		0.080	03:37
Air Blank		0.000	03:38
Subject Sample	# 1	0.085	03:38
Breath Volume		2.730	03:38
Air Blank		0.000	03:39
Air Blank		0.000	03:41
Subject Sample	# 2	0.084	03:41
Breath Volume		2.042	03:41
Air Blank		0.000	03:42
Control Test		0.079	03:42
Air Blank		0.000	03:43
Diagnostics Check		OK	03:43

Data Download Date/Time: 11/17/2017 15:27

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARI,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SENTMANAT,A
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 10/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:25

SUBJECT'S LAST NAME: ALCOCER BECERRA

DATE/TIME OF TEST: 11/12/2017 05:03:10

Diagnostics Check		OK	04:56
Air Blank		0.000	04:57
Control Test		0.079	04:57
Air Blank		0.000	04:58
Subject Sample	# 1	0.195	04:58
Breath Volume		2.917	04:58
Air Blank		0.000	04:59
Air Blank		0.000	05:01
Subject Sample	# 2	0.195	05:01
Breath Volume		1.617	05:01
Air Blank		0.000	05:02
Control Test		0.078	05:02
Air Blank		0.000	05:03
Diagnostics Check		OK	05:03

Data Download Date/Time: 11/17/2017 15:28

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAIN,JEANETTE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LEVEY,SCOTT
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 10/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: MARQUEZ

DATE/TIME OF TEST: 11/14/2017 01:14:21

Diagnostics Check		OK	01:07
Air Blank		0.000	01:08
Control Test		0.080	01:08
Air Blank		0.000	01:09
Subject Sample	# 1	0.000	01:09
Breath Volume		1.285	01:09
Air Blank		0.000	01:10
Air Blank		0.000	01:12
Subject Sample	# 2	0.000	01:12
Breath Volume		1.414	01:12
Air Blank		0.000	01:13
Control Test		0.080	01:13
Air Blank		0.000	01:14
Diagnostics Check		OK	01:14

Data Download Date/Time: 11/17/2017 15:28

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAIN,JEANETTE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: GUARIN,EDGAR
ARRESTING OFFICER AGENCY: Lantana PD
LAST AGENCY INSPECTION: 10/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 16:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:53

SUBJECT'S LAST NAME: HOGARTY

DATE/TIME OF TEST: 11/14/2017 18:24:04

Diagnostics Check		OK	18:17
Air Blank		0.000	18:18
Control Test		0.080	18:18
Air Blank		0.000	18:18
Subject Sample	# 1	0.000	18:19
Breath Volume		2.878	18:19
Air Blank		0.000	18:20
Air Blank		0.000	18:22
Subject Sample	# 2	0.000	18:22
Breath Volume		2.425	18:22
Air Blank		0.000	18:23
Control Test		0.080	18:23
Air Blank		0.000	18:23
Diagnostics Check		OK	18:24

Data Download Date/Time: 11/17/2017 15:28

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAIN,JEANETTE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WALTON,THOMAS
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 10/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:15

SUBJECT'S LAST NAME: BAIN

DATE/TIME OF TEST: 11/14/2017 21:41:53

Diagnostics Check		OK	21:38	
Air Blank		0.000	21:38	
Control Test		0.080	21:39	
Air Blank		0.000	21:39	
Subject Sample	# 1		21:40	Subject Test Refused
Breath Volume		0.000	21:40	
Air Blank		0.000	21:40	
Control Test		0.079	21:41	
Air Blank		0.000	21:41	
Diagnostics Check		OK	21:41	

Data Download Date/Time: 11/17/2017 15:28

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAIN,JEANETTE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ANNESE,MORRIS
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 10/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 11/15/2017 23:15:05

Diagnostics Check		OK	23:08
Air Blank		0.000	23:08
Control Test		0.080	23:08
Air Blank		0.000	23:09
Subject Sample	# 1	0.205	23:10
Breath Volume		1.433	23:10
Air Blank		0.000	23:11
Air Blank		0.000	23:13
Subject Sample	# 2	0.206	23:13
Breath Volume		1.265	23:13
Air Blank		0.000	23:14
Control Test		0.078	23:14
Air Blank		0.000	23:15
Diagnostics Check		OK	23:15

Data Download Date/Time: 11/17/2017 15:28

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: HERNANDEZ,MARTIN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 10/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: PONCE

DATE/TIME OF TEST: 11/16/2017 02:08:38

Diagnostics Check		OK	02:00
Air Blank		0.000	02:01
Control Test		0.079	02:01
Air Blank		0.000	02:01
Subject Sample	# 1	0.157	02:03
Breath Volume		2.175	02:03
Air Blank		0.000	02:04
Air Blank		0.000	02:06
Subject Sample	# 2	0.162	02:07
Breath Volume		1.871	02:07
Air Blank		0.000	02:07
Control Test		0.079	02:08
Air Blank		0.000	02:08
Diagnostics Check		OK	02:08

Data Download Date/Time: 11/17/2017 15:28

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: POINTU,PATRICK
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 10/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:07

SUBJECT'S LAST NAME: BRYAN

DATE/TIME OF TEST: 11/16/2017 04:36:18

Diagnostics Check		OK	04:33	
Air Blank		0.000	04:33	
Control Test		0.079	04:34	
Air Blank		0.000	04:34	
Subject Sample	# 1		04:34	Subject Test Refused
Breath Volume		0.000	04:34	
Air Blank		0.000	04:35	
Control Test		0.080	04:35	
Air Blank		0.000	04:36	
Diagnostics Check		OK	04:36	

Data Download Date/Time: 11/17/2017 15:28

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: YOCHUM,CRAIG
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 10/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:25

SUBJECT'S LAST NAME: CHUTA

DATE/TIME OF TEST: 11/16/2017 05:57:39

Diagnostics Check		OK	05:50
Air Blank		0.000	05:51
Control Test		0.080	05:51
Air Blank		0.000	05:51
Subject Sample	# 1	0.245	05:52
Breath Volume		2.031	05:52
Air Blank		0.000	05:53
Air Blank		0.000	05:55
Subject Sample	# 2	0.241	05:55
Breath Volume		2.449	05:55
Air Blank		0.000	05:56
Control Test		0.079	05:57
Air Blank		0.000	05:57
Diagnostics Check		OK	05:57

Data Download Date/Time: 11/17/2017 15:28

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAIN,JEANETTE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LEVEY,SCOTT
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 10/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:25

SUBJECT'S LAST NAME: KNOX

DATE/TIME OF TEST: 11/16/2017 20:54:34

Diagnostics Check		OK	20:47
Air Blank		0.000	20:48
Control Test		0.080	20:48
Air Blank		0.000	20:49
Subject Sample	# 1	0.217	20:49
Breath Volume		1.347	20:49
Air Blank		0.000	20:50
Air Blank		0.000	20:52
Subject Sample	# 2	0.217	20:52
Breath Volume		2.253	20:52
Air Blank		0.000	20:53
Control Test		0.079	20:53
Air Blank		0.000	20:54
Diagnostics Check		OK	20:54

Data Download Date/Time: 11/17/2017 15:28

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAIN,JEANETTE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITE,E. KEITH
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 10/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:35

SUBJECT'S LAST NAME: RICHARD

DATE/TIME OF TEST: 11/16/2017 21:24:51

Diagnostics Check		OK	21:16
Air Blank		0.000	21:17
Control Test		0.080	21:17
Air Blank		0.000	21:18
Subject Sample	# 1	0.000	21:19
Breath Volume		1.535	21:19
Air Blank		0.000	21:19
Air Blank		0.000	21:21
Subject Sample	# 2	0.000	21:23
Breath Volume		1.335	21:23
Air Blank		0.000	21:23
Control Test		0.080	21:24
Air Blank		0.000	21:24
Diagnostics Check		OK	21:24

Data Download Date/Time: 11/17/2017 15:28

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAIN,JEANETTE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: HERNANDEZ,MARTIN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 10/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:27

SUBJECT'S LAST NAME: WADE

DATE/TIME OF TEST: 11/16/2017 22:20:43

Diagnostics Check		OK	22:05	
Air Blank		0.000	22:05	
Control Test		0.079	22:06	
Air Blank		0.000	22:06	
Subject Sample	# 1	0.146	22:09	Volume Not Met
Breath Volume		0.152	22:09	
Air Blank		0.000	22:10	
Air Blank		0.000	22:12	
Subject Sample	# 2	0.169	22:13	
Breath Volume		1.300	22:13	
Air Blank		0.000	22:14	
Air Blank		0.000	22:15	
Subject Sample	# 3	0.164	22:19	
Breath Volume		1.109	22:19	
Air Blank		0.000	22:19	
Control Test		0.079	22:20	
Air Blank		0.000	22:20	
Diagnostics Check		OK	22:20	

Data Download Date/Time: 11/17/2017 15:28

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARI,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: TAVARES,LUCAS
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 10/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: VELIS

DATE/TIME OF TEST: 11/17/2017 01:22:34

Diagnostics Check		OK	01:16
Air Blank		0.000	01:16
Control Test		0.079	01:17
Air Blank		0.000	01:17
Subject Sample	# 1	0.247	01:18
Breath Volume		1.246	01:18
Air Blank		0.000	01:18
Air Blank		0.000	01:20
Subject Sample	# 2	0.234	01:20
Breath Volume		1.296	01:20
Air Blank		0.000	01:21
Control Test		0.079	01:21
Air Blank		0.000	01:22
Diagnostics Check		OK	01:22

Data Download Date/Time: 11/17/2017 15:28

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARI,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SCHNEIDER,JOSHUA
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 10/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: GRANT

DATE/TIME OF TEST: 11/17/2017 02:09:15

Diagnostics Check		OK	02:03
Air Blank		0.000	02:03
Control Test		0.079	02:03
Air Blank		0.000	02:04
Subject Sample	# 1	0.119	02:04
Breath Volume		2.125	02:04
Air Blank		0.000	02:05
Air Blank		0.000	02:07
Subject Sample	# 2	0.112	02:07
Breath Volume		3.582	02:07
Air Blank		0.000	02:08
Control Test		0.079	02:08
Air Blank		0.000	02:09
Diagnostics Check		OK	02:09

Data Download Date/Time: 11/17/2017 15:28

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARLEY,JOSEPH,E
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: FARLEY,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 11/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:56

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 11/02/2017 02:23:59

Diagnostics Check		OK	02:21	
Air Blank		0.000	02:21	
Control Test		0.079	02:21	
Air Blank		0.000	02:22	
Subject Sample	# 1		02:22	Subject Test Refused
Breath Volume		0.000	02:22	
Air Blank		0.000	02:23	
Control Test		0.080	02:23	
Air Blank		0.000	02:23	
Diagnostics Check		OK	02:23	

Data Download Date/Time: 11/23/2017 21:20

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARLEY,JOSEPH,E
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: FARLEY,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 11/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:38

SUBJECT'S LAST NAME: HORTON

DATE/TIME OF TEST: 11/09/2017 22:06:51

Diagnostics Check		OK	21:58	
Air Blank		0.000	21:59	
Control Test		0.082	21:59	
Air Blank		0.000	22:00	
Subject Sample	# 1	0.261	22:01	
Breath Volume		1.632	22:01	
Air Blank		0.000	22:02	
Air Blank		0.000	22:03	
Subject Sample	# 2	0.257	22:04	
Breath Volume		3.085	22:04	
Air Blank		0.000	22:05	
Control Test		0.081	22:05	
Air Blank		0.000	22:06	
Diagnostics Check		Fail	22:06	Analytical Stability Test
Air Blank		0.000	22:06	

Data Download Date/Time: 11/23/2017 21:20

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006262

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANSARD,JOSHUA,D
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: FEELEY,JAMES
ARRESTING OFFICER AGENCY: Ormond Beach PD
LAST AGENCY INSPECTION: 10/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 03:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:13

SUBJECT'S LAST NAME: VILLARS II

DATE/TIME OF TEST: 11/01/2017 03:46:46

Diagnostics Check		OK	03:40
Air Blank		0.000	03:40
Control Test		0.080	03:41
Air Blank		0.000	03:41
Subject Sample	# 1	0.251	03:42
Breath Volume		2.921	03:42
Air Blank		0.000	03:42
Air Blank		0.000	03:44
Subject Sample	# 2	0.239	03:45
Breath Volume		2.312	03:45
Air Blank		0.000	03:45
Control Test		0.080	03:46
Air Blank		0.000	03:46
Diagnostics Check		OK	03:46

Data Download Date/Time: 11/06/2017 13:30

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006262

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARMINGTON,WILLIAM,G
BREATH TEST OPERATOR AGENCY: Ormond Beach PD
ARRESTING OFFICER: DOVINE,JOHN
ARRESTING OFFICER AGENCY: Ormond Beach PD
LAST AGENCY INSPECTION: 10/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 20:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: HENDERSON JR

DATE/TIME OF TEST: 11/01/2017 21:52:29

Diagnostics Check		OK	21:45
Air Blank		0.000	21:46
Control Test		0.081	21:46
Air Blank		0.000	21:47
Subject Sample	# 1	0.000	21:47
Breath Volume		1.964	21:47
Air Blank		0.000	21:47
Air Blank		0.000	21:49
Subject Sample	# 2	0.000	21:51
Breath Volume		1.460	21:51
Air Blank		0.000	21:51
Control Test		0.081	21:51
Air Blank		0.000	21:52
Diagnostics Check		OK	21:52

Data Download Date/Time: 11/06/2017 13:30

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006262

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,ANDREW,J
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: WEAVER,WILLIAM
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 11/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 20:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:12

SUBJECT'S LAST NAME: COFFIN

DATE/TIME OF TEST: 11/25/2017 21:42:46

Diagnostics Check		OK	21:36
Air Blank		0.000	21:36
Control Test		0.080	21:37
Air Blank		0.000	21:37
Subject Sample	# 1	0.290	21:38
Breath Volume		1.261	21:38
Air Blank		0.000	21:38
Air Blank		0.000	21:40
Subject Sample	# 2	0.295	21:41
Breath Volume		1.847	21:41
Air Blank		0.000	21:41
Control Test		0.080	21:42
Air Blank		0.000	21:42
Diagnostics Check		OK	21:42

Data Download Date/Time: 12/28/2017 11:52

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006384

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VISMAN,CLAYTON,A
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: VISMAN,CLAYTON
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 10/13/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: WRIGHT

DATE/TIME OF TEST: 11/02/2017 03:21:51

Diagnostics Check		OK	03:14	
Air Blank		0.000	03:14	
Control Test		0.080	03:14	
Air Blank		0.000	03:15	
Subject Sample	# 1		03:18	No Sample Provided
Breath Volume		0.000	03:18	
Air Blank		0.000	03:18	
Air Blank		0.000	03:20	
Subject Sample	# 2		03:21	Subject Test Refused
Breath Volume		0.000	03:21	
Air Blank			03:21	RFI Detect
Air Blank		0.000	03:21	

Data Download Date/Time: 11/16/2017 21:42

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006384

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEMEN,AARON,M
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: VISMAN,CLAYTON
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 10/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: WRIGHT

DATE/TIME OF TEST: 11/02/2017 03:37:07

Diagnostics Check		OK	03:25	
Air Blank		0.000	03:26	
Control Test		0.081	03:26	
Air Blank		0.000	03:26	
Subject Sample	# 1		03:30	No Sample Provided
Breath Volume		0.000	03:30	
Air Blank		0.000	03:30	
Air Blank		0.000	03:32	
Subject Sample	# 2		03:35	No Sample Provided
Breath Volume		0.000	03:35	
Air Blank		0.000	03:36	
Control Test		0.081	03:36	
Air Blank		0.000	03:37	
Diagnostics Check		OK	03:37	

Data Download Date/Time: 11/16/2017 21:42

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006384

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEMEN,AARON,M
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: N/A,N/A
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 10/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: ZERO TOL
ARREST TIME: 00:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:46

SUBJECT'S LAST NAME: SHEERIN

DATE/TIME OF TEST: 11/03/2017 01:15:34

Diagnostics Check		OK	01:08
Air Blank		0.000	01:09
Control Test		0.080	01:09
Air Blank		0.000	01:10
Subject Sample	# 1	0.065	01:10
Breath Volume		2.101	01:10
Air Blank		0.000	01:11
Air Blank		0.000	01:13
Subject Sample	# 2	0.063	01:13
Breath Volume		1.777	01:13
Air Blank		0.000	01:14
Control Test		0.080	01:14
Air Blank		0.000	01:15
Diagnostics Check		OK	01:15

Data Download Date/Time: 11/16/2017 21:43

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006384

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEMEN,AARON,M
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: STEMEN,AARON
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 10/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: BOCKMAN

DATE/TIME OF TEST: 11/03/2017 03:33:05

Diagnostics Check		OK	03:25
Air Blank		0.000	03:26
Control Test		0.080	03:26
Air Blank		0.000	03:27
Subject Sample	# 1	0.161	03:28
Breath Volume		3.902	03:28
Air Blank		0.000	03:28
Air Blank		0.000	03:30
Subject Sample	# 2	0.156	03:31
Breath Volume		2.132	03:31
Air Blank		0.000	03:32
Control Test		0.080	03:32
Air Blank		0.000	03:32
Diagnostics Check		OK	03:33

Data Download Date/Time: 11/16/2017 21:43

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006384

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEMEN,AARON,M
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: STEMEN,AARON
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 10/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:02

SUBJECT'S LAST NAME: CALAF VAZQUEZ

DATE/TIME OF TEST: 11/04/2017 03:29:16

Diagnostics Check		OK	03:22
Air Blank		0.000	03:22
Control Test		0.080	03:23
Air Blank		0.000	03:23
Subject Sample	# 1	0.228	03:24
Breath Volume		1.687	03:24
Air Blank		0.000	03:25
Air Blank		0.000	03:26
Subject Sample	# 2	0.225	03:27
Breath Volume		2.070	03:27
Air Blank		0.000	03:28
Control Test		0.080	03:28
Air Blank		0.000	03:29
Diagnostics Check		OK	03:29

Data Download Date/Time: 11/16/2017 21:43

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006384

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEMEN,AARON,M
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: STEMEN,AARON
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 10/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:23

SUBJECT'S LAST NAME: MARLOWE

DATE/TIME OF TEST: 11/05/2017 00:49:37

Diagnostics Check		OK	00:43
Air Blank		0.000	00:43
Control Test		0.080	00:43
Air Blank		0.000	00:44
Subject Sample	# 1	0.000	00:45
Breath Volume		4.449	00:45
Air Blank		0.000	00:45
Air Blank		0.000	00:47
Subject Sample	# 2	0.000	00:48
Breath Volume		2.968	00:48
Air Blank		0.000	00:48
Control Test		0.080	00:48
Air Blank		0.000	00:49
Diagnostics Check		OK	00:49

Data Download Date/Time: 11/16/2017 21:43

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006384

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEMEN,AARON,M
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: STEMEN,AARON
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 10/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:23

SUBJECT'S LAST NAME: JACOBS

DATE/TIME OF TEST: 11/09/2017 22:53:09

Diagnostics Check		OK	22:43	
Air Blank		0.000	22:44	
Control Test		0.080	22:44	
Air Blank		0.000	22:45	
Subject Sample	# 1		22:48	No Sample Provided
Breath Volume		0.000	22:48	
Air Blank		0.000	22:48	
Air Blank		0.000	22:50	
Subject Sample	# 2		22:51	Subject Test Refused
Breath Volume		0.000	22:51	
Air Blank		0.000	22:52	
Control Test		0.079	22:52	
Air Blank		0.000	22:53	
Diagnostics Check		OK	22:53	

Data Download Date/Time: 11/16/2017 21:43

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006384

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEMEN,AARON,M
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: STEMEN,AARON
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 10/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:46

SUBJECT'S LAST NAME: SHEFFLER

DATE/TIME OF TEST: 11/11/2017 02:27:59

Diagnostics Check		OK	02:16	
Air Blank		0.000	02:16	
Control Test		0.080	02:17	
Air Blank		0.000	02:17	
Subject Sample	# 1		02:20	No Sample Provided
Breath Volume		0.000	02:20	
Air Blank		0.000	02:21	
Air Blank		0.000	02:23	
Subject Sample	# 2		02:26	No Sample Provided
Breath Volume		0.000	02:26	
Air Blank		0.000	02:26	
Control Test		0.082	02:27	
Air Blank		0.000	02:27	
Diagnostics Check		OK	02:27	

Data Download Date/Time: 11/16/2017 21:43

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006384

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEMEN,AARON,M
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: STEMEN,AARON
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 11/16/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:59

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 11/17/2017 00:27:40

Diagnostics Check		OK	00:21
Air Blank		0.000	00:21
Control Test		0.081	00:21
Air Blank		0.000	00:22
Subject Sample	# 1	0.140	00:23
Breath Volume		1.574	00:23
Air Blank		0.000	00:23
Air Blank		0.000	00:25
Subject Sample	# 2	0.141	00:26
Breath Volume		1.300	00:26
Air Blank		0.000	00:26
Control Test		0.080	00:27
Air Blank		0.000	00:27
Diagnostics Check		OK	00:27

Data Download Date/Time: 12/06/2017 20:31

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006384

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KEIMIG,EMORIE,I
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: KEIMIG,EMORIE
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 11/16/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:59

SUBJECT'S LAST NAME: BEADE

DATE/TIME OF TEST: 11/18/2017 21:29:12

Diagnostics Check		OK	21:21
Air Blank		0.000	21:22
Control Test		0.081	21:22
Air Blank		0.000	21:23
Subject Sample	# 1	0.212	21:24
Breath Volume		1.796	21:24
Air Blank		0.000	21:25
Air Blank		0.000	21:26
Subject Sample	# 2	0.196	21:27
Breath Volume		1.437	21:27
Air Blank		0.000	21:28
Control Test		0.079	21:28
Air Blank		0.000	21:29
Diagnostics Check		OK	21:29

Data Download Date/Time: 12/06/2017 20:31

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006384

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEMEN,AARON,M
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: STEMEN,AARON
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 11/16/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:44

SUBJECT'S LAST NAME: VIOLA

DATE/TIME OF TEST: 11/30/2017 04:14:34

Diagnostics Check		OK	04:07
Air Blank		0.000	04:08
Control Test		0.080	04:08
Air Blank		0.000	04:08
Subject Sample	# 1	0.157	04:09
Breath Volume		1.847	04:09
Air Blank		0.000	04:10
Air Blank		0.000	04:12
Subject Sample	# 2	0.163	04:12
Breath Volume		1.964	04:12
Air Blank		0.000	04:13
Control Test		0.080	04:13
Air Blank		0.000	04:14
Diagnostics Check		OK	04:14

Data Download Date/Time: 12/06/2017 20:31

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PECYLAK,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:44

SUBJECT'S LAST NAME: SANTIZ GOMEZ

DATE/TIME OF TEST: 11/01/2017 03:11:24

Diagnostics Check		OK	03:04
Air Blank		0.000	03:05
Control Test		0.081	03:05
Air Blank		0.000	03:05
Subject Sample	# 1	0.210	03:06
Breath Volume		1.484	03:06
Air Blank		0.000	03:07
Air Blank		0.000	03:09
Subject Sample	# 2	0.215	03:09
Breath Volume		1.570	03:09
Air Blank		0.000	03:10
Control Test		0.080	03:10
Air Blank		0.000	03:11
Diagnostics Check		OK	03:11

Data Download Date/Time: 11/21/2017 14:17

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 10/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:49

SUBJECT'S LAST NAME: VELAZQUEZ

DATE/TIME OF TEST: 11/02/2017 04:16:18

Diagnostics Check		OK	04:09
Air Blank		0.000	04:09
Control Test		0.080	04:10
Air Blank		0.000	04:10
Subject Sample	# 1	0.141	04:11
Breath Volume		1.355	04:11
Air Blank		0.000	04:12
Air Blank		0.000	04:14
Subject Sample	# 2	0.137	04:14
Breath Volume		1.636	04:14
Air Blank		0.000	04:15
Control Test		0.079	04:15
Air Blank		0.000	04:16
Diagnostics Check		OK	04:16

Data Download Date/Time: 11/21/2017 14:17

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GONZALEZ,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 10/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:46

SUBJECT'S LAST NAME: LIGHTNER

DATE/TIME OF TEST: 11/03/2017 03:14:06

Diagnostics Check		OK	03:06
Air Blank		0.000	03:06
Control Test		0.080	03:07
Air Blank		0.000	03:07
Subject Sample	# 1	0.085	03:08
Breath Volume		2.000	03:08
Air Blank		0.000	03:09
Air Blank		0.000	03:11
Subject Sample	# 2	0.090	03:12
Breath Volume		1.984	03:12
Air Blank		0.000	03:13
Control Test		0.080	03:13
Air Blank		0.000	03:14
Diagnostics Check		OK	03:14

Data Download Date/Time: 11/21/2017 14:17

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FRANKLIN,R
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:54

SUBJECT'S LAST NAME: CAMPBELL

DATE/TIME OF TEST: 11/03/2017 04:21:00

Diagnostics Check		OK	04:14
Air Blank		0.000	04:15
Control Test		0.081	04:15
Air Blank		0.000	04:16
Subject Sample	# 1	0.158	04:16
Breath Volume		2.066	04:16
Air Blank		0.000	04:17
Air Blank		0.000	04:18
Subject Sample	# 2	0.156	04:19
Breath Volume		2.496	04:19
Air Blank		0.000	04:20
Control Test		0.079	04:20
Air Blank		0.000	04:20
Diagnostics Check		OK	04:21

Data Download Date/Time: 11/21/2017 14:17

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ORNDORFF,W
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: ROBBIO IV

DATE/TIME OF TEST: 11/04/2017 00:56:55

Diagnostics Check		OK	00:50
Air Blank		0.000	00:50
Control Test		0.081	00:50
Air Blank		0.000	00:51
Subject Sample	# 1	0.218	00:52
Breath Volume		1.695	00:52
Air Blank		0.000	00:52
Air Blank		0.000	00:54
Subject Sample	# 2	0.226	00:55
Breath Volume		2.097	00:55
Air Blank		0.000	00:55
Control Test		0.080	00:56
Air Blank		0.000	00:56
Diagnostics Check		OK	00:56

Data Download Date/Time: 11/21/2017 14:17

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ANDERSON,C
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 10/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:15

SUBJECT'S LAST NAME: CUERVO

DATE/TIME OF TEST: 11/04/2017 05:43:52

Diagnostics Check		OK	05:35	
Air Blank		0.000	05:35	
Control Test		0.080	05:36	
Air Blank		0.000	05:36	
Subject Sample	# 1		05:39	No Sample Provided
Breath Volume		0.000	05:39	
Air Blank		0.000	05:40	
Air Blank		0.000	05:42	
Subject Sample	# 2		05:42	Subject Test Refused
Breath Volume		0.000	05:42	
Air Blank		0.000	05:42	
Control Test		0.081	05:43	
Air Blank		0.000	05:43	
Diagnostics Check		OK	05:43	

Data Download Date/Time: 11/21/2017 14:17

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BLANCHARD,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 10/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:08

SUBJECT'S LAST NAME: VILLEGAS

DATE/TIME OF TEST: 11/05/2017 01:35:37

Diagnostics Check		OK	01:28
Air Blank		0.000	01:28
Control Test		0.080	01:28
Air Blank		0.000	01:29
Subject Sample	# 1	0.197	01:30
Breath Volume		1.195	01:30
Air Blank		0.000	01:31
Air Blank		0.000	01:32
Subject Sample	# 2	0.197	01:33
Breath Volume		1.214	01:33
Air Blank		0.000	01:34
Control Test		0.079	01:34
Air Blank		0.000	01:35
Diagnostics Check		OK	01:35

Data Download Date/Time: 11/21/2017 14:17

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: NIEUWENDAM,R
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 10/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:27

SUBJECT'S LAST NAME: GOMEZ

DATE/TIME OF TEST: 11/05/2017 02:56:04

Diagnostics Check		OK	02:49
Air Blank		0.000	02:49
Control Test		0.081	02:50
Air Blank		0.000	02:50
Subject Sample	# 1	0.057	02:51
Breath Volume		2.953	02:51
Air Blank		0.000	02:52
Air Blank		0.000	02:53
Subject Sample	# 2	0.056	02:54
Breath Volume		3.398	02:54
Air Blank		0.000	02:55
Control Test		0.079	02:55
Air Blank		0.000	02:55
Diagnostics Check		OK	02:56

Data Download Date/Time: 11/21/2017 14:17

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SIMPKINS,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:13

SUBJECT'S LAST NAME: TRAMEL JR

DATE/TIME OF TEST: 11/05/2017 05:36:11

Diagnostics Check		OK	05:33	
Air Blank		0.000	05:33	
Control Test		0.080	05:33	
Air Blank		0.000	05:34	
Subject Sample	# 1		05:34	Subject Test Refused
Breath Volume		0.000	05:34	
Air Blank		0.000	05:35	
Control Test		0.080	05:35	
Air Blank		0.000	05:36	
Diagnostics Check		OK	05:36	

Data Download Date/Time: 11/21/2017 14:17

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VAN-TREESE,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 10/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:39

SUBJECT'S LAST NAME: SALIAJ

DATE/TIME OF TEST: 11/06/2017 03:07:04

Diagnostics Check		OK	02:59
Air Blank		0.000	02:59
Control Test		0.080	03:00
Air Blank		0.000	03:00
Subject Sample	# 1	0.158	03:02
Breath Volume		1.937	03:02
Air Blank		0.000	03:02
Air Blank		0.000	03:04
Subject Sample	# 2	0.164	03:05
Breath Volume		2.265	03:05
Air Blank		0.000	03:06
Control Test		0.079	03:06
Air Blank		0.000	03:06
Diagnostics Check		OK	03:07

Data Download Date/Time: 11/21/2017 14:17

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ORNDORFF,W
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:57

SUBJECT'S LAST NAME: LEE

DATE/TIME OF TEST: 11/06/2017 04:25:12

Diagnostics Check		OK	04:17
Air Blank		0.000	04:17
Control Test		0.080	04:18
Air Blank		0.000	04:18
Subject Sample	# 1	0.192	04:20
Breath Volume		1.675	04:20
Air Blank		0.000	04:20
Air Blank		0.000	04:22
Subject Sample	# 2	0.172	04:23
Breath Volume		1.265	04:23
Air Blank		0.000	04:24
Control Test		0.079	04:24
Air Blank		0.000	04:25
Diagnostics Check		OK	04:25

Data Download Date/Time: 11/21/2017 14:17

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:26

SUBJECT'S LAST NAME: MERIDA ROBLERO

DATE/TIME OF TEST: 11/09/2017 01:53:15

Diagnostics Check		OK	01:46
Air Blank		0.000	01:47
Control Test		0.080	01:47
Air Blank		0.000	01:48
Subject Sample	# 1	0.204	01:48
Breath Volume		2.191	01:48
Air Blank		0.000	01:49
Air Blank		0.000	01:51
Subject Sample	# 2	0.203	01:51
Breath Volume		1.394	01:51
Air Blank		0.000	01:52
Control Test		0.080	01:52
Air Blank		0.000	01:53
Diagnostics Check		OK	01:53

Data Download Date/Time: 11/21/2017 14:17

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SHEA,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:03

SUBJECT'S LAST NAME: GOINS

DATE/TIME OF TEST: 11/09/2017 21:38:43

Diagnostics Check		OK	21:23	
Air Blank		0.000	21:23	
Control Test		0.080	21:23	
Air Blank		0.000	21:24	
Subject Sample	# 1	0.180	21:27	Volume Not Met
Breath Volume		0.195	21:27	
Air Blank		0.000	21:28	
Air Blank		0.000	21:30	
Subject Sample	# 2	0.219	21:32	
Breath Volume		3.003	21:32	
Air Blank		0.000	21:33	
Air Blank		0.000	21:35	
Subject Sample	# 3	0.207	21:37	
Breath Volume		2.097	21:37	
Air Blank		0.000	21:37	
Control Test		0.077	21:38	
Air Blank		0.000	21:38	
Diagnostics Check		OK	21:38	

Data Download Date/Time: 11/21/2017 14:17

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRENNAN,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:28

SUBJECT'S LAST NAME: AHRENS

DATE/TIME OF TEST: 11/11/2017 01:56:08

Diagnostics Check		OK	01:48
Air Blank		0.000	01:49
Control Test		0.080	01:49
Air Blank		0.000	01:50
Subject Sample	# 1	0.137	01:51
Breath Volume		1.398	01:51
Air Blank		0.000	01:52
Air Blank		0.000	01:53
Subject Sample	# 2	0.139	01:54
Breath Volume		1.714	01:54
Air Blank		0.000	01:55
Control Test		0.080	01:55
Air Blank		0.000	01:56
Diagnostics Check		OK	01:56

Data Download Date/Time: 11/21/2017 14:17

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HUTTON,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 10/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:32

SUBJECT'S LAST NAME: DUHART

DATE/TIME OF TEST: 11/11/2017 03:57:03

Diagnostics Check		OK	03:52	
Air Blank		0.000	03:53	
Control Test		0.080	03:53	
Air Blank		0.000	03:53	
Subject Sample	# 1		03:55	Subject Test Refused
Breath Volume		0.000	03:55	
Air Blank		0.000	03:56	
Control Test		0.081	03:56	
Air Blank		0.000	03:56	
Diagnostics Check		OK	03:57	

Data Download Date/Time: 11/21/2017 14:17

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FRANKLIN,R
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:58

SUBJECT'S LAST NAME: MORRIS

DATE/TIME OF TEST: 11/12/2017 00:25:50

Diagnostics Check		OK	00:19
Air Blank		0.000	00:19
Control Test		0.081	00:19
Air Blank		0.000	00:20
Subject Sample	# 1	0.131	00:21
Breath Volume		2.429	00:21
Air Blank		0.000	00:21
Air Blank		0.000	00:23
Subject Sample	# 2	0.132	00:24
Breath Volume		1.531	00:24
Air Blank		0.000	00:24
Control Test		0.080	00:25
Air Blank		0.000	00:25
Diagnostics Check		OK	00:25

Data Download Date/Time: 11/21/2017 14:17

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRIMBLE,TERESA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GONZALEZ,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:35

SUBJECT'S LAST NAME: LOGAN

DATE/TIME OF TEST: 11/12/2017 05:03:32

Diagnostics Check		OK	04:55
Air Blank		0.000	04:55
Control Test		0.080	04:56
Air Blank		0.000	04:56
Subject Sample	# 1	0.164	04:58
Breath Volume		1.597	04:58
Air Blank		0.000	04:59
Air Blank		0.000	05:01
Subject Sample	# 2	0.165	05:01
Breath Volume		1.558	05:01
Air Blank		0.000	05:02
Control Test		0.078	05:02
Air Blank		0.000	05:03
Diagnostics Check		OK	05:03

Data Download Date/Time: 11/21/2017 14:17

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FRANKLIN,R
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:59

SUBJECT'S LAST NAME: ANTONIO SANJUAN

DATE/TIME OF TEST: 11/12/2017 05:27:02

Diagnostics Check		OK	05:19
Air Blank		0.000	05:20
Control Test		0.080	05:20
Air Blank		0.000	05:21
Subject Sample	# 1	0.188	05:22
Breath Volume		2.941	05:22
Air Blank		0.000	05:23
Air Blank		0.000	05:24
Subject Sample	# 2	0.194	05:25
Breath Volume		3.136	05:25
Air Blank		0.000	05:26
Control Test		0.079	05:26
Air Blank		0.000	05:26
Diagnostics Check		OK	05:27

Data Download Date/Time: 11/21/2017 14:17

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FRANKLIN,R
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:41

SUBJECT'S LAST NAME: MAUND

DATE/TIME OF TEST: 11/16/2017 05:13:27

Diagnostics Check		OK	05:01	
Air Blank		0.000	05:01	
Control Test		0.080	05:01	
Air Blank		0.000	05:02	
Subject Sample	# 1	0.127	05:05	Volume Not Met
Breath Volume		1.019	05:05	
Air Blank		0.000	05:06	
Air Blank		0.000	05:08	
Subject Sample	# 2	0.128	05:08	
Breath Volume		2.367	05:08	
Air Blank		0.000	05:09	
Air Blank		0.000	05:11	
Subject Sample	# 3	0.129	05:11	
Breath Volume		2.027	05:11	
Air Blank		0.000	05:12	
Control Test		0.080	05:12	
Air Blank		0.000	05:13	
Diagnostics Check		OK	05:13	

Data Download Date/Time: 11/21/2017 14:17

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: NOBLE,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 10/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:29

SUBJECT'S LAST NAME: CARDOSO

DATE/TIME OF TEST: 11/17/2017 02:56:15

Diagnostics Check		OK	02:50
Air Blank		0.000	02:50
Control Test		0.081	02:50
Air Blank		0.000	02:51
Subject Sample	# 1	0.000	02:51
Breath Volume		3.324	02:51
Air Blank		0.000	02:52
Air Blank		0.000	02:54
Subject Sample	# 2	0.000	02:54
Breath Volume		3.613	02:54
Air Blank		0.000	02:55
Control Test		0.080	02:55
Air Blank		0.000	02:56
Diagnostics Check		OK	02:56

Data Download Date/Time: 11/21/2017 14:17

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ORNDORFF,W
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:01

SUBJECT'S LAST NAME: PINILLA

DATE/TIME OF TEST: 11/18/2017 00:29:44

Diagnostics Check		OK	00:21
Air Blank		0.000	00:21
Control Test		0.081	00:22
Air Blank		0.000	00:22
Subject Sample	# 1	0.143	00:24
Breath Volume		2.152	00:24
Air Blank		0.000	00:25
Air Blank		0.000	00:27
Subject Sample	# 2	0.145	00:28
Breath Volume		1.105	00:28
Air Blank		0.000	00:28
Control Test		0.079	00:29
Air Blank		0.000	00:29
Diagnostics Check		OK	00:29

Data Download Date/Time: 11/21/2017 14:17

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: COSTA,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 11/19/2017 00:22:34

Diagnostics Check		OK	00:10	
Air Blank		0.000	00:11	
Control Test		0.080	00:11	
Air Blank		0.000	00:12	
Subject Sample	# 1	0.135	00:15	Volume Not Met
Breath Volume		0.066	00:15	
Air Blank		0.000	00:16	
Air Blank		0.000	00:17	
Subject Sample	# 2	0.108	00:20	Volume Not Met
Breath Volume		0.390	00:20	
Air Blank		0.000	00:21	
Control Test		0.078	00:21	
Air Blank		0.000	00:22	
Diagnostics Check		OK	00:22	

Data Download Date/Time: 11/21/2017 14:17

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SIMPKINS,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:55

SUBJECT'S LAST NAME: TEHUINTLE-CALIHUA

DATE/TIME OF TEST: 11/19/2017 05:23:49

Diagnostics Check		OK	05:16
Air Blank		0.000	05:17
Control Test		0.080	05:17
Air Blank		0.000	05:17
Subject Sample	# 1	0.158	05:18
Breath Volume		1.949	05:18
Air Blank		0.000	05:19
Air Blank		0.000	05:21
Subject Sample	# 2	0.156	05:22
Breath Volume		1.203	05:22
Air Blank		0.000	05:22
Control Test		0.079	05:23
Air Blank		0.000	05:23
Diagnostics Check		OK	05:23

Data Download Date/Time: 11/21/2017 14:17

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WOODEN,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 18:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:45

SUBJECT'S LAST NAME: KELL

DATE/TIME OF TEST: 11/19/2017 20:09:30

Diagnostics Check		OK	20:05	
Air Blank		0.000	20:06	
Control Test		0.080	20:06	
Air Blank		0.000	20:07	
Subject Sample	# 1		20:08	Subject Test Refused
Breath Volume		0.000	20:08	
Air Blank		0.000	20:08	
Control Test		0.081	20:08	
Air Blank		0.000	20:09	
Diagnostics Check		OK	20:09	

Data Download Date/Time: 11/21/2017 14:17

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GONZALEZ,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:52

SUBJECT'S LAST NAME: CACERES

DATE/TIME OF TEST: 11/21/2017 05:19:58

Diagnostics Check		OK	05:12
Air Blank		0.000	05:12
Control Test		0.081	05:13
Air Blank		0.000	05:13
Subject Sample	# 1	0.152	05:15
Breath Volume		2.179	05:15
Air Blank		0.000	05:16
Air Blank		0.000	05:17
Subject Sample	# 2	0.152	05:18
Breath Volume		2.058	05:18
Air Blank		0.000	05:19
Control Test		0.079	05:19
Air Blank		0.000	05:19
Diagnostics Check		OK	05:19

Data Download Date/Time: 11/21/2017 14:17

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006386

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NORWOOD,CHRISTOPHER,J
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: CORONILLA,DANIEL
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 11/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 437738
DRY GAS EXPIRY DATE: 06/08/2019
VIOLATION CODE: DUI
ARREST TIME: 23:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:06

SUBJECT'S LAST NAME: ORTEGA

DATE/TIME OF TEST: 11/30/2017 00:50:06

Diagnostics Check		OK	00:43
Air Blank		0.000	00:44
Control Test		0.082	00:44
Air Blank		0.000	00:45
Subject Sample	# 1	0.124	00:45
Breath Volume		1.859	00:45
Air Blank		0.000	00:46
Air Blank		0.000	00:48
Subject Sample	# 2	0.120	00:48
Breath Volume		1.476	00:48
Air Blank		0.000	00:49
Control Test		0.082	00:49
Air Blank		0.000	00:50
Diagnostics Check		OK	00:50

Data Download Date/Time: 12/26/2017 18:58

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEITE,STEVEN,G
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: LEITE,STEVEN
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 10/16/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:46

SUBJECT'S LAST NAME: DOSS

DATE/TIME OF TEST: 11/04/2017 02:18:37

Diagnostics Check		OK	02:06	
Air Blank		0.000	02:06	
Control Test		0.081	02:07	
Air Blank		0.000	02:07	
Subject Sample	# 1	0.146	02:10	Slope Not Level
Breath Volume		2.101	02:10	
Air Blank		0.000	02:11	
Air Blank		0.000	02:13	
Subject Sample	# 2	0.150	02:13	
Breath Volume		2.210	02:13	
Air Blank		0.000	02:14	
Air Blank		0.000	02:16	
Subject Sample	# 3	0.149	02:17	
Breath Volume		2.191	02:17	
Air Blank		0.000	02:17	
Control Test		0.079	02:18	
Air Blank		0.000	02:18	
Diagnostics Check		OK	02:18	

Data Download Date/Time: 11/29/2017 09:22

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENDRIX,TIMOTHY,D
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: HENDRIX,TIMOTHY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 10/16/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:11

SUBJECT'S LAST NAME: MACDOUGALL

DATE/TIME OF TEST: 11/06/2017 20:53:10

Diagnostics Check		OK	20:46
Air Blank		0.000	20:46
Control Test		0.081	20:47
Air Blank		0.000	20:47
Subject Sample	# 1	0.200	20:48
Breath Volume		2.769	20:48
Air Blank		0.000	20:48
Air Blank		0.000	20:50
Subject Sample	# 2	0.201	20:51
Breath Volume		2.214	20:51
Air Blank		0.000	20:52
Control Test		0.079	20:52
Air Blank		0.000	20:53
Diagnostics Check		OK	20:53

Data Download Date/Time: 11/29/2017 09:22

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENLEY,WILLIAM,C
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: SECOR,BROC
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 10/16/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:49

SUBJECT'S LAST NAME: FOSTER

DATE/TIME OF TEST: 11/08/2017 22:16:19

Diagnostics Check		OK	22:09
Air Blank		0.000	22:10
Control Test		0.080	22:10
Air Blank		0.000	22:10
Subject Sample	# 1	0.225	22:11
Breath Volume		2.003	22:11
Air Blank		0.000	22:12
Air Blank		0.000	22:14
Subject Sample	# 2	0.217	22:14
Breath Volume		1.605	22:14
Air Blank		0.000	22:15
Control Test		0.080	22:15
Air Blank		0.000	22:16
Diagnostics Check		OK	22:16

Data Download Date/Time: 11/29/2017 09:22

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENLEY,WILLIAM,C
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: HAMPTON,LOGAN
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 10/16/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:29

SUBJECT'S LAST NAME: CREMEANS

DATE/TIME OF TEST: 11/08/2017 22:56:28

Diagnostics Check		OK	22:49
Air Blank		0.000	22:50
Control Test		0.081	22:50
Air Blank		0.000	22:51
Subject Sample	# 1	0.094	22:51
Breath Volume		3.324	22:51
Air Blank		0.000	22:52
Air Blank		0.000	22:54
Subject Sample	# 2	0.085	22:54
Breath Volume		2.246	22:54
Air Blank		0.000	22:55
Control Test		0.081	22:55
Air Blank		0.000	22:56
Diagnostics Check		OK	22:56

Data Download Date/Time: 11/29/2017 09:22

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEEKS,JEREMIAH,S
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: MEEKS,JEREMIAH
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 10/16/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: BLACKMON

DATE/TIME OF TEST: 11/14/2017 01:59:37

Diagnostics Check		OK	01:53
Air Blank		0.000	01:53
Control Test		0.081	01:54
Air Blank		0.000	01:54
Subject Sample	# 1	0.271	01:55
Breath Volume		2.289	01:55
Air Blank		0.000	01:55
Air Blank		0.000	01:57
Subject Sample	# 2	0.281	01:58
Breath Volume		2.035	01:58
Air Blank		0.000	01:58
Control Test		0.078	01:59
Air Blank		0.000	01:59
Diagnostics Check		OK	01:59

Data Download Date/Time: 11/29/2017 09:22

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEEKS,JEREMIAH,S
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: ASHTON,DAVID
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 10/16/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:53

SUBJECT'S LAST NAME: NORMAN

DATE/TIME OF TEST: 11/18/2017 00:23:31

Diagnostics Check		OK	00:20	
Air Blank		0.000	00:21	
Control Test		0.082	00:21	
Air Blank		0.000	00:21	
Subject Sample	# 1		00:22	Subject Test Refused
Breath Volume		0.000	00:22	
Air Blank		0.000	00:22	
Control Test		0.081	00:22	
Air Blank		0.000	00:23	
Diagnostics Check		OK	00:23	

Data Download Date/Time: 11/29/2017 09:22

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEEKS,JEREMIAH,S
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: GRAVES,TIMOTHY
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 10/16/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:33

SUBJECT'S LAST NAME: FORD-HUMBLES

DATE/TIME OF TEST: 11/18/2017 19:59:12

Diagnostics Check		OK	19:53
Air Blank		0.000	19:53
Control Test		0.081	19:53
Air Blank		0.000	19:54
Subject Sample	# 1	0.000	19:54
Breath Volume		2.417	19:54
Air Blank		0.000	19:55
Air Blank		0.000	19:57
Subject Sample	# 2	0.000	19:57
Breath Volume		2.562	19:57
Air Blank		0.000	19:58
Control Test		0.080	19:58
Air Blank		0.000	19:59
Diagnostics Check		OK	19:59

Data Download Date/Time: 11/29/2017 09:22

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEEKS,JEREMIAH,S
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: GRAVES,TIMOTHY
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 10/16/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:29

SUBJECT'S LAST NAME: PRIDGEN

DATE/TIME OF TEST: 11/19/2017 01:56:38

Diagnostics Check		OK	01:49
Air Blank		0.000	01:50
Control Test		0.081	01:50
Air Blank		0.000	01:51
Subject Sample	# 1	0.197	01:51
Breath Volume		1.199	01:51
Air Blank		0.000	01:52
Air Blank		0.000	01:54
Subject Sample	# 2	0.205	01:54
Breath Volume		1.613	01:54
Air Blank		0.000	01:55
Control Test		0.079	01:56
Air Blank		0.000	01:56
Diagnostics Check		OK	01:56

Data Download Date/Time: 11/29/2017 09:22

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HYDER,CHRIS,M
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: PHILLIPS,REECE
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 10/16/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:37

SUBJECT'S LAST NAME: KIMBROUGH

DATE/TIME OF TEST: 11/25/2017 23:24:38

Diagnostics Check		OK	23:18
Air Blank		0.000	23:18
Control Test		0.082	23:19
Air Blank		0.000	23:19
Subject Sample	# 1	0.110	23:19
Breath Volume		1.214	23:19
Air Blank		0.000	23:20
Air Blank		0.000	23:22
Subject Sample	# 2	0.109	23:23
Breath Volume		1.937	23:23
Air Blank		0.000	23:23
Control Test		0.081	23:24
Air Blank		0.000	23:24
Diagnostics Check		OK	23:24

Data Download Date/Time: 11/29/2017 09:22

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TADLOCK,J,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: TADLOCK,J
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 10/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 05:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:10

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 11/04/2017 06:34:07

Diagnostics Check		OK	06:31	
Air Blank		0.000	06:31	
Control Test		0.079	06:31	
Air Blank		0.000	06:32	
Subject Sample	# 1		06:32	Subject Test Refused
Breath Volume		0.000	06:32	
Air Blank		0.000	06:33	
Control Test		0.079	06:33	
Air Blank		0.000	06:34	
Diagnostics Check		OK	06:34	

Data Download Date/Time: 11/30/2017 14:20

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VASQUEZ,JASON,A
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: RIOS,RICHARD
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 10/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 20:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:25

SUBJECT'S LAST NAME: SOLIS

DATE/TIME OF TEST: 11/04/2017 21:04:13

Diagnostics Check		OK	20:51	
Air Blank		0.000	20:52	
Control Test		0.078	20:52	
Air Blank		0.000	20:53	
Subject Sample	# 1	0.133	20:56	Volume Not Met
Breath Volume		0.718	20:56	
Air Blank		0.000	20:57	
Air Blank		0.000	20:58	
Subject Sample	# 2	0.159	20:59	
Breath Volume		1.847	20:59	
Air Blank		0.000	21:00	
Air Blank		0.000	21:02	
Subject Sample	# 3	0.164	21:02	
Breath Volume		2.050	21:02	
Air Blank		0.000	21:03	
Control Test		0.076	21:03	
Air Blank		0.000	21:04	
Diagnostics Check		OK	21:04	

Data Download Date/Time: 11/30/2017 14:20

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VILLARREAL,FABIANA,C
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: VILLARREAL,FABIANA
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 10/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 11/05/2017 03:17:48

Diagnostics Check		OK	03:11
Air Blank		0.000	03:11
Control Test		0.079	03:12
Air Blank		0.000	03:12
Subject Sample	# 1	0.153	03:13
Breath Volume		2.152	03:13
Air Blank		0.000	03:13
Air Blank		0.000	03:15
Subject Sample	# 2	0.149	03:16
Breath Volume		3.445	03:16
Air Blank		0.000	03:16
Control Test		0.079	03:17
Air Blank		0.000	03:17
Diagnostics Check		OK	03:17

Data Download Date/Time: 11/30/2017 14:20

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERNANDEZ,VANESSA,G
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: GUTIERREZ,JESSICA
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 10/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 19:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: GARMENDIA

DATE/TIME OF TEST: 11/05/2017 20:28:46

Diagnostics Check		OK	20:21
Air Blank		0.000	20:22
Control Test		0.078	20:22
Air Blank		0.000	20:23
Subject Sample	# 1	0.133	20:24
Breath Volume		2.175	20:24
Air Blank		0.000	20:24
Air Blank		0.000	20:26
Subject Sample	# 2	0.135	20:27
Breath Volume		1.644	20:27
Air Blank		0.000	20:27
Control Test		0.079	20:28
Air Blank		0.000	20:28
Diagnostics Check		OK	20:28

Data Download Date/Time: 11/30/2017 14:20

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MENA,ALEXANDER,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: DE PRADO,DANIEL
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 10/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:43

SUBJECT'S LAST NAME: MILANES ACOSTA

DATE/TIME OF TEST: 11/06/2017 02:13:46

Diagnostics Check		OK	02:03
Air Blank		0.000	02:04
Control Test		0.080	02:04
Air Blank		0.000	02:05
Subject Sample	# 1	0.213	02:06
Breath Volume		1.484	02:06
Air Blank		0.000	02:07
Air Blank		0.000	02:09
Subject Sample	# 2	0.218	02:11
Breath Volume		1.378	02:11
Air Blank		0.000	02:12
Control Test		0.077	02:13
Air Blank		0.000	02:13
Diagnostics Check		OK	02:13

Data Download Date/Time: 11/30/2017 14:20

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CABRAL,EDUARDO,
BREATH TEST OPERATOR AGENCY: Coral Gables PD
ARRESTING OFFICER: CABRAL,EDUARDO
ARRESTING OFFICER AGENCY: Coral Gables PD
LAST AGENCY INSPECTION: 10/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 17:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:45

SUBJECT'S LAST NAME: CAMPOS GRINON

DATE/TIME OF TEST: 11/08/2017 19:16:18

Diagnostics Check		OK	19:06	
Air Blank		0.000	19:06	
Control Test		0.079	19:07	
Air Blank		0.000	19:07	
Subject Sample	# 1	0.081	19:10	Volume Not Met
Breath Volume		0.082	19:10	
Air Blank		0.000	19:11	
Air Blank		0.000	19:13	
Subject Sample	# 2		19:14	Subject Test Refused
Breath Volume		0.113	19:14	
Air Blank		0.000	19:15	
Control Test		0.077	19:15	
Air Blank		0.000	19:16	
Diagnostics Check		OK	19:16	

Data Download Date/Time: 11/30/2017 14:20

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,ALBERTO,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: RODRIGUEZ,ALBERTO
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 10/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 19:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:06

SUBJECT'S LAST NAME: MATUTE ORELLANA

DATE/TIME OF TEST: 11/09/2017 20:36:50

Diagnostics Check		OK	20:30
Air Blank		0.000	20:30
Control Test		0.079	20:30
Air Blank		0.000	20:31
Subject Sample	# 1	0.233	20:31
Breath Volume		2.675	20:31
Air Blank		0.000	20:32
Air Blank		0.000	20:34
Subject Sample	# 2	0.231	20:35
Breath Volume		1.234	20:35
Air Blank		0.000	20:35
Control Test		0.077	20:36
Air Blank		0.000	20:36
Diagnostics Check		OK	20:36

Data Download Date/Time: 11/30/2017 14:20

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MENA,ALEXANDER,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: TARGIA,FRANK
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 10/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 00:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:23

SUBJECT'S LAST NAME: RAMIREZ

DATE/TIME OF TEST: 11/16/2017 00:49:28

Diagnostics Check		OK	00:43	
Air Blank		0.000	00:44	
Control Test		0.080	00:44	
Air Blank		0.000	00:45	
Subject Sample	# 1		00:47	Subject Test Refused
Breath Volume		0.183	00:47	
Air Blank		0.000	00:48	
Control Test		0.079	00:48	
Air Blank		0.000	00:49	
Diagnostics Check		OK	00:49	

Data Download Date/Time: 11/30/2017 14:20

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILES,STEPHANIE,N
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: SILES,STEPHANIE
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 10/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 00:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: CASTELLON

DATE/TIME OF TEST: 11/17/2017 01:22:28

Diagnostics Check		OK	01:15
Air Blank		0.000	01:16
Control Test		0.079	01:16
Air Blank		0.000	01:16
Subject Sample	# 1	0.222	01:17
Breath Volume		2.953	01:17
Air Blank		0.000	01:18
Air Blank		0.000	01:20
Subject Sample	# 2	0.224	01:20
Breath Volume		2.039	01:20
Air Blank		0.000	01:21
Control Test		0.078	01:21
Air Blank		0.000	01:22
Diagnostics Check		OK	01:22

Data Download Date/Time: 11/30/2017 14:20

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,ALEXANDER,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: GARCIA,ALEXANDER
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 10/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 01:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: VARGA

DATE/TIME OF TEST: 11/17/2017 05:22:22

Diagnostics Check		OK	05:19	
Air Blank		0.000	05:19	
Control Test		0.079	05:20	
Air Blank		0.000	05:20	
Subject Sample	# 1		05:21	Subject Test Refused
Breath Volume		0.000	05:21	
Air Blank		0.000	05:21	
Control Test		0.078	05:21	
Air Blank		0.000	05:22	
Diagnostics Check		OK	05:22	

Data Download Date/Time: 11/30/2017 14:20

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAMIREZ,MOISES,A
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: CARBONO,DOUGLAS
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 10/26/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:37

SUBJECT'S LAST NAME: SANCHEZ

DATE/TIME OF TEST: 11/18/2017 03:09:10

Diagnostics Check		OK	03:00	
Air Blank		0.000	03:00	
Control Test		0.079	03:01	
Air Blank		0.000	03:01	
Subject Sample	# 1	0.139	03:04	
Breath Volume		2.957	03:04	
Air Blank		0.000	03:04	
Air Blank		0.000	03:06	
Subject Sample	# 2		03:08	Sequence Aborted
Breath Volume		0.000	03:08	
Air Blank		0.000	03:09	

Data Download Date/Time: 11/30/2017 14:20

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAMIREZ,MOISES,A
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: CARBONO,DOUGLAS
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 10/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:09

SUBJECT'S LAST NAME: SANCHEZ

DATE/TIME OF TEST: 11/18/2017 03:36:45

Diagnostics Check		OK	03:30
Air Blank		0.000	03:30
Control Test		0.079	03:31
Air Blank		0.000	03:31
Subject Sample	# 1	0.135	03:32
Breath Volume		2.265	03:32
Air Blank		0.000	03:32
Air Blank		0.000	03:34
Subject Sample	# 2	0.136	03:35
Breath Volume		2.664	03:35
Air Blank		0.000	03:35
Control Test		0.079	03:36
Air Blank		0.000	03:36
Diagnostics Check		OK	03:36

Data Download Date/Time: 11/30/2017 14:20

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MENA,ALEXANDER,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: KENHAN,MAX
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 10/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 23:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: LORENZO PADILLA

DATE/TIME OF TEST: 11/19/2017 00:10:41

Diagnostics Check		OK	00:06	
Air Blank		0.000	00:06	
Control Test		0.079	00:07	
Air Blank		0.000	00:07	
Subject Sample	# 1		00:09	Subject Test Refused
Breath Volume		0.000	00:09	
Air Blank		0.000	00:09	
Control Test		0.079	00:10	
Air Blank		0.000	00:10	
Diagnostics Check		OK	00:10	

Data Download Date/Time: 11/30/2017 14:20

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,SANDRA,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: HERNANDEZ,JASIEL
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 10/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 23:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:07

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 11/22/2017 00:36:22

Diagnostics Check		OK	00:28	
Air Blank		0.000	00:28	
Control Test		0.080	00:28	
Air Blank		0.000	00:29	
Subject Sample	# 1		00:32	No Sample Provided
Breath Volume		0.000	00:32	
Air Blank		0.000	00:33	
Air Blank		0.000	00:35	
Subject Sample	# 2		00:35	RFI Detect
Breath Volume		0.000	00:35	
Air Blank			00:35	RFI Detect
Air Blank		0.000	00:36	

Data Download Date/Time: 11/30/2017 14:20

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,JASIEL,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: HERNANDEZ,JASIEL
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 10/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 23:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:08

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 11/22/2017 00:56:15

Diagnostics Check		OK	00:41	
Air Blank		0.000	00:41	
Control Test		0.079	00:42	
Air Blank		0.000	00:42	
Subject Sample	# 1	0.204	00:45	Volume Not Met
Breath Volume		0.738	00:45	
Air Blank		0.000	00:46	
Air Blank		0.000	00:48	
Subject Sample	# 2	0.238	00:50	
Breath Volume		2.015	00:50	
Air Blank		0.000	00:51	
Air Blank		0.000	00:52	
Subject Sample	# 3	0.246	00:54	
Breath Volume		1.722	00:54	
Air Blank		0.000	00:55	
Control Test		0.078	00:55	
Air Blank		0.000	00:56	
Diagnostics Check		OK	00:56	

Data Download Date/Time: 11/30/2017 14:20

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,ALBERTO,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: RODRIGUEZ,ALBERTO
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 10/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: COURT ORD
ARREST TIME: 13:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:23

SUBJECT'S LAST NAME: MCDONOUGH

DATE/TIME OF TEST: 11/25/2017 13:51:32

Diagnostics Check		OK	13:45
Air Blank		0.000	13:45
Control Test		0.079	13:46
Air Blank		0.000	13:46
Subject Sample	# 1	0.000	13:47
Breath Volume		1.359	13:47
Air Blank		0.000	13:47
Air Blank		0.000	13:49
Subject Sample	# 2	0.000	13:50
Breath Volume		1.820	13:50
Air Blank		0.000	13:50
Control Test		0.077	13:50
Air Blank		0.000	13:51
Diagnostics Check		OK	13:51

Data Download Date/Time: 11/30/2017 14:20

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,CRISTINO,
BREATH TEST OPERATOR AGENCY: Coral Gables PD
ARRESTING OFFICER: PUGA,JORGE
ARRESTING OFFICER AGENCY: Coral Gables PD
LAST AGENCY INSPECTION: 10/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 17:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:44

SUBJECT'S LAST NAME: SANCHEZ

DATE/TIME OF TEST: 11/28/2017 20:11:18

Diagnostics Check		OK	20:04
Air Blank		0.000	20:05
Control Test		0.079	20:05
Air Blank		0.000	20:05
Subject Sample	# 1	0.064	20:06
Breath Volume		2.988	20:06
Air Blank		0.000	20:07
Air Blank		0.000	20:08
Subject Sample	# 2	0.064	20:09
Breath Volume		2.250	20:09
Air Blank		0.000	20:10
Control Test		0.079	20:10
Air Blank		0.000	20:11
Diagnostics Check		OK	20:11

Data Download Date/Time: 11/30/2017 14:20

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HEREIDA RUBIO,ROBERTO,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: HEREIDA RUBIO,ROBERTO
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 10/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 21:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:08

SUBJECT'S LAST NAME: MUNOZ

DATE/TIME OF TEST: 11/09/2017 22:35:20

Diagnostics Check		OK	22:28
Air Blank		0.000	22:29
Control Test		0.080	22:29
Air Blank		0.000	22:29
Subject Sample	# 1	0.000	22:30
Breath Volume		2.140	22:30
Air Blank		0.000	22:31
Air Blank		0.000	22:33
Subject Sample	# 2	0.000	22:33
Breath Volume		1.355	22:33
Air Blank		0.000	22:34
Control Test		0.080	22:34
Air Blank		0.000	22:35
Diagnostics Check		OK	22:35

Data Download Date/Time: 11/30/2017 15:46

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,JONATHAN,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: HERNANDEZ,JASIEL
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 10/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 01:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:41

SUBJECT'S LAST NAME: RAMIREZ

DATE/TIME OF TEST: 11/17/2017 02:08:11

Diagnostics Check		OK	02:02
Air Blank		0.000	02:02
Control Test		0.080	02:02
Air Blank		0.000	02:03
Subject Sample	# 1	0.000	02:03
Breath Volume		2.554	02:03
Air Blank		0.000	02:04
Air Blank		0.000	02:06
Subject Sample	# 2	0.000	02:06
Breath Volume		1.546	02:06
Air Blank		0.000	02:07
Control Test		0.079	02:07
Air Blank		0.000	02:08
Diagnostics Check		OK	02:08

Data Download Date/Time: 11/30/2017 15:46

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006473

INSTRUMENT REGISTERED TO: PINECREST PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRUZ,EDISON,B
BREATH TEST OPERATOR AGENCY: Village of Pinecrest
ARRESTING OFFICER: GARCIA,JOSE
ARRESTING OFFICER AGENCY: Village of Pinecrest
LAST AGENCY INSPECTION: 10/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A6
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 07:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:13

SUBJECT'S LAST NAME: VICUNA

DATE/TIME OF TEST: 11/04/2017 08:43:10

Diagnostics Check		OK	08:33
Air Blank		0.000	08:33
Control Test		0.078	08:33
Air Blank		0.000	08:34
Subject Sample	# 1	0.118	08:35
Breath Volume		1.601	08:35
Air Blank		0.000	08:36
Air Blank		0.000	08:38
Subject Sample	# 2	0.124	08:41
Breath Volume		1.562	08:41
Air Blank		0.000	08:42
Control Test		0.078	08:42
Air Blank		0.000	08:43
Diagnostics Check		OK	08:43

Data Download Date/Time: 11/22/2017 09:14

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006473

INSTRUMENT REGISTERED TO: PINECREST PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRUZ,EDISON,B
BREATH TEST OPERATOR AGENCY: Village of Pinecrest
ARRESTING OFFICER: CRUZ,EDISON
ARRESTING OFFICER AGENCY: Village of Pinecrest
LAST AGENCY INSPECTION: 11/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A6
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:29

SUBJECT'S LAST NAME: MEINKING

DATE/TIME OF TEST: 11/27/2017 21:05:35

Diagnostics Check		OK	20:50	
Air Blank		0.000	20:51	
Control Test		0.078	20:51	
Air Blank		0.000	20:52	
Subject Sample	# 1	0.000	20:55	
Breath Volume		1.472	20:55	
Air Blank		0.000	20:55	
Air Blank		0.000	20:57	
Subject Sample	# 2	0.000	21:00	Volume Not Met
Breath Volume		0.347	21:00	
Air Blank		0.000	21:01	
Air Blank		0.000	21:03	
Subject Sample	# 3	0.000	21:04	
Breath Volume		2.117	21:04	
Air Blank		0.000	21:04	
Control Test		0.077	21:04	
Air Blank		0.000	21:05	
Diagnostics Check		OK	21:05	

Data Download Date/Time: 12/21/2017 10:03

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SERPHOS,STEVEN,A
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SERPHOS,STEVEN
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 10/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A3
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:49

SUBJECT'S LAST NAME: HOLMES

DATE/TIME OF TEST: 11/05/2017 22:18:43

Diagnostics Check		OK	22:10	
Air Blank		0.000	22:10	
Control Test		0.076	22:11	
Air Blank		0.000	22:11	
Subject Sample	# 1	0.148	22:14	
Breath Volume		2.398	22:14	
Air Blank		0.000	22:14	
Air Blank		0.000	22:16	
Subject Sample	# 2	0.133	22:17	
Breath Volume		1.277	22:17	
Air Blank		0.000	22:17	
Control Test		0.074	22:18	Control Outside Tolerance
Air Blank		0.000	22:18	

Data Download Date/Time: 11/06/2017 14:39

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SERPHOS,STEVEN,A
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SERPHOS,STEVEN
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 10/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A3
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:49

SUBJECT'S LAST NAME: HOLMES

DATE/TIME OF TEST: 11/05/2017 22:31:08

Diagnostics Check		OK	22:21
Air Blank		0.000	22:21
Control Test		0.076	22:22
Air Blank		0.000	22:22
Subject Sample	# 1	0.140	22:25
Breath Volume		1.761	22:25
Air Blank		0.000	22:25
Air Blank		0.000	22:27
Subject Sample	# 2	0.129	22:29
Breath Volume		1.421	22:29
Air Blank		0.000	22:30
Control Test		0.075	22:30
Air Blank		0.000	22:31
Diagnostics Check		OK	22:31

Data Download Date/Time: 11/06/2017 14:39

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SERPHOS,STEVEN,A
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SERPHOS,STEVEN
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 11/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:47

SUBJECT'S LAST NAME: TIMA

DATE/TIME OF TEST: 11/16/2017 00:17:04

Diagnostics Check		OK	00:08
Air Blank		0.000	00:08
Control Test		0.077	00:09
Air Blank		0.000	00:09
Subject Sample	# 1	0.192	00:11
Breath Volume		2.078	00:11
Air Blank		0.000	00:12
Air Blank		0.000	00:14
Subject Sample	# 2	0.179	00:15
Breath Volume		1.386	00:15
Air Blank		0.000	00:16
Control Test		0.076	00:16
Air Blank		0.000	00:16
Diagnostics Check		OK	00:17

Data Download Date/Time: 12/13/2017 17:16

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SERPHOS,STEVEN,A
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SERPHOS,STEVEN
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 11/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:22

SUBJECT'S LAST NAME: RASCHKE

DATE/TIME OF TEST: 11/20/2017 03:59:42

Diagnostics Check		OK	03:46	
Air Blank		0.000	03:47	
Control Test		0.077	03:47	
Air Blank		0.000	03:47	
Subject Sample	# 1	0.166	03:51	Volume Not Met
Breath Volume		0.019	03:51	
Air Blank		0.000	03:51	
Air Blank		0.000	03:53	
Subject Sample	# 2	0.183	03:55	
Breath Volume		1.753	03:55	
Air Blank		0.000	03:55	
Air Blank		0.000	03:57	
Subject Sample	# 3	0.176	03:58	
Breath Volume		2.347	03:58	
Air Blank		0.000	03:58	
Control Test		0.075	03:59	
Air Blank		0.000	03:59	
Diagnostics Check		OK	03:59	

Data Download Date/Time: 12/13/2017 17:16

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SERPHOS,STEVEN,A
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SERPHOS,STEVEN
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 11/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:25

SUBJECT'S LAST NAME: STREAMS

DATE/TIME OF TEST: 11/20/2017 22:52:01

Diagnostics Check		OK	22:45
Air Blank		0.000	22:46
Control Test		0.076	22:46
Air Blank		0.000	22:47
Subject Sample	# 1	0.186	22:47
Breath Volume		2.757	22:47
Air Blank		0.000	22:48
Air Blank		0.000	22:49
Subject Sample	# 2	0.192	22:50
Breath Volume		3.355	22:50
Air Blank		0.000	22:51
Control Test		0.076	22:51
Air Blank		0.000	22:51
Diagnostics Check		OK	22:52

Data Download Date/Time: 12/13/2017 17:16

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SERPHOS,STEVEN,A
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SERPHOS,STEVEN
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 11/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:58

SUBJECT'S LAST NAME: CAREY

DATE/TIME OF TEST: 11/26/2017 21:26:55

Diagnostics Check		OK	21:20
Air Blank		0.000	21:20
Control Test		0.078	21:20
Air Blank		0.000	21:21
Subject Sample	# 1	0.104	21:21
Breath Volume		2.800	21:21
Air Blank		0.000	21:22
Air Blank		0.000	21:24
Subject Sample	# 2	0.099	21:25
Breath Volume		2.121	21:25
Air Blank		0.000	21:25
Control Test		0.079	21:26
Air Blank		0.000	21:26
Diagnostics Check		OK	21:26

Data Download Date/Time: 12/13/2017 17:16

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALMER,SAMANTHA,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: POINTU,PATRICK
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 10/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:56

SUBJECT'S LAST NAME: HALL

DATE/TIME OF TEST: 11/01/2017 04:26:44

Diagnostics Check		OK	04:19
Air Blank		0.000	04:20
Control Test		0.080	04:20
Air Blank		0.000	04:21
Subject Sample	# 1	0.135	04:22
Breath Volume		2.597	04:22
Air Blank		0.000	04:22
Air Blank		0.000	04:24
Subject Sample	# 2	0.129	04:25
Breath Volume		2.597	04:25
Air Blank		0.000	04:25
Control Test		0.079	04:26
Air Blank		0.000	04:26
Diagnostics Check		OK	04:26

Data Download Date/Time: 11/17/2017 13:00

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: GONZALEZ,JOSH
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 10/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: HUSULIAK

DATE/TIME OF TEST: 11/02/2017 01:55:33

Diagnostics Check		OK	01:49
Air Blank		0.000	01:49
Control Test		0.081	01:50
Air Blank		0.000	01:50
Subject Sample	# 1	0.130	01:51
Breath Volume		3.695	01:51
Air Blank		0.000	01:51
Air Blank		0.000	01:53
Subject Sample	# 2	0.130	01:53
Breath Volume		3.699	01:53
Air Blank		0.000	01:54
Control Test		0.080	01:54
Air Blank		0.000	01:55
Diagnostics Check		OK	01:55

Data Download Date/Time: 11/17/2017 13:00

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SECORD,RYAN
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 10/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 06:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:15

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 11/03/2017 07:50:35

Diagnostics Check		OK	07:44
Air Blank		0.000	07:44
Control Test		0.080	07:45
Air Blank		0.000	07:45
Subject Sample	# 1	0.127	07:46
Breath Volume		1.355	07:46
Air Blank		0.000	07:46
Air Blank		0.000	07:48
Subject Sample	# 2	0.128	07:48
Breath Volume		1.523	07:48
Air Blank		0.000	07:49
Control Test		0.080	07:49
Air Blank		0.000	07:50
Diagnostics Check		OK	07:50

Data Download Date/Time: 11/17/2017 13:00

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LEVEY,SCOTT
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 10/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: CALAVETTA

DATE/TIME OF TEST: 11/04/2017 03:34:04

Diagnostics Check		OK	03:18	
Air Blank		0.000	03:19	
Control Test		0.081	03:19	
Air Blank		0.000	03:20	
Subject Sample	# 1	0.166	03:23	
Breath Volume		1.121	03:23	
Air Blank		0.000	03:24	
Air Blank		0.000	03:25	
Subject Sample	# 2	0.141	03:28	Volume Not Met
Breath Volume		0.875	03:28	
Air Blank		0.000	03:29	
Air Blank		0.000	03:31	
Subject Sample	# 3	0.159	03:32	
Breath Volume		1.308	03:32	
Air Blank		0.000	03:33	
Control Test		0.077	03:33	
Air Blank		0.000	03:33	
Diagnostics Check		OK	03:34	

Data Download Date/Time: 11/17/2017 13:00

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MARSEILLE,ANGELO
ARRESTING OFFICER AGENCY: Delray Beach PD
LAST AGENCY INSPECTION: 10/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:56

SUBJECT'S LAST NAME: SILLER

DATE/TIME OF TEST: 11/04/2017 22:26:58

Diagnostics Check		OK	22:20
Air Blank		0.000	22:21
Control Test		0.080	22:21
Air Blank		0.000	22:21
Subject Sample	# 1	0.270	22:22
Breath Volume		2.972	22:22
Air Blank		0.000	22:23
Air Blank		0.000	22:24
Subject Sample	# 2	0.266	22:25
Breath Volume		3.125	22:25
Air Blank		0.000	22:26
Control Test		0.079	22:26
Air Blank		0.000	22:26
Diagnostics Check		OK	22:26

Data Download Date/Time: 11/17/2017 13:00

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: IV,ANDREW
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 10/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:03

SUBJECT'S LAST NAME: MONDESIR

DATE/TIME OF TEST: 11/05/2017 03:37:01

Diagnostics Check		OK	03:29
Air Blank		0.000	03:30
Control Test		0.080	03:30
Air Blank		0.000	03:31
Subject Sample	# 1	0.166	03:32
Breath Volume		1.621	03:32
Air Blank		0.000	03:32
Air Blank		0.000	03:34
Subject Sample	# 2	0.172	03:35
Breath Volume		1.335	03:35
Air Blank		0.000	03:36
Control Test		0.080	03:36
Air Blank		0.000	03:36
Diagnostics Check		OK	03:37

Data Download Date/Time: 11/17/2017 13:00

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: YOUNG,JOHNNY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 10/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:17

SUBJECT'S LAST NAME: SULLIVAN

DATE/TIME OF TEST: 11/05/2017 04:50:56

Diagnostics Check		OK	04:41
Air Blank		0.000	04:42
Control Test		0.081	04:42
Air Blank		0.000	04:43
Subject Sample	# 1	0.123	04:45
Breath Volume		1.164	04:45
Air Blank		0.000	04:46
Air Blank		0.000	04:48
Subject Sample	# 2	0.124	04:49
Breath Volume		1.437	04:49
Air Blank		0.000	04:50
Control Test		0.079	04:50
Air Blank		0.000	04:50
Diagnostics Check		OK	04:50

Data Download Date/Time: 11/17/2017 13:00

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: GUBISH,MATTHEW
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 10/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 07:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:31

SUBJECT'S LAST NAME: VILLALBA

DATE/TIME OF TEST: 11/05/2017 08:03:49

Diagnostics Check		OK	07:57
Air Blank		0.000	07:57
Control Test		0.080	07:58
Air Blank		0.000	07:58
Subject Sample	# 1	0.136	07:59
Breath Volume		1.230	07:59
Air Blank		0.000	07:59
Air Blank		0.000	08:01
Subject Sample	# 2	0.145	08:02
Breath Volume		1.265	08:02
Air Blank		0.000	08:02
Control Test		0.080	08:03
Air Blank		0.000	08:03
Diagnostics Check		OK	08:03

Data Download Date/Time: 11/17/2017 13:00

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALMER,SAMANTHA,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SCHAEFER,JOHN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 10/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:32

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 11/06/2017 01:03:42

Diagnostics Check		OK	00:57
Air Blank		0.000	00:57
Control Test		0.080	00:57
Air Blank		0.000	00:58
Subject Sample	# 1	0.180	00:58
Breath Volume		1.218	00:58
Air Blank		0.000	00:59
Air Blank		0.000	01:01
Subject Sample	# 2	0.182	01:02
Breath Volume		1.960	01:02
Air Blank		0.000	01:02
Control Test		0.079	01:03
Air Blank		0.000	01:03
Diagnostics Check		OK	01:03

Data Download Date/Time: 11/17/2017 13:00

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: IV,ANDREW
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 10/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:36

SUBJECT'S LAST NAME: VASQUEZ-LOPEZ

DATE/TIME OF TEST: 11/07/2017 02:07:26

Diagnostics Check		OK	02:00
Air Blank		0.000	02:01
Control Test		0.080	02:01
Air Blank		0.000	02:02
Subject Sample	# 1	0.123	02:02
Breath Volume		2.640	02:02
Air Blank		0.000	02:03
Air Blank		0.000	02:05
Subject Sample	# 2	0.123	02:05
Breath Volume		2.910	02:05
Air Blank		0.000	02:06
Control Test		0.080	02:06
Air Blank		0.000	02:07
Diagnostics Check		OK	02:07

Data Download Date/Time: 11/17/2017 13:00

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALMER,SAMANTHA,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: DOMINGUEZ,JAYMES
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 10/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 07:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:00

SUBJECT'S LAST NAME: GILLET

DATE/TIME OF TEST: 11/08/2017 08:36:27

Diagnostics Check		OK	08:29
Air Blank		0.000	08:30
Control Test		0.080	08:30
Air Blank		0.000	08:31
Subject Sample	# 1	0.000	08:31
Breath Volume		2.375	08:31
Air Blank		0.000	08:32
Air Blank		0.000	08:34
Subject Sample	# 2	0.000	08:35
Breath Volume		2.328	08:35
Air Blank		0.000	08:35
Control Test		0.080	08:35
Air Blank		0.000	08:36
Diagnostics Check		OK	08:36

Data Download Date/Time: 11/17/2017 13:00

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MITCHELL,CEDRIC
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 10/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: ELDER

DATE/TIME OF TEST: 11/09/2017 00:07:21

Diagnostics Check		OK	00:01
Air Blank		0.000	00:01
Control Test		0.080	00:02
Air Blank		0.000	00:02
Subject Sample	# 1	0.074	00:02
Breath Volume		3.449	00:02
Air Blank		0.000	00:03
Air Blank		0.000	00:05
Subject Sample	# 2	0.077	00:05
Breath Volume		3.472	00:05
Air Blank		0.000	00:06
Control Test		0.080	00:06
Air Blank		0.000	00:07
Diagnostics Check		OK	00:07

Data Download Date/Time: 11/17/2017 13:00

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WALTON,THOMAS
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 10/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: JOYA

DATE/TIME OF TEST: 11/10/2017 00:40:52

Diagnostics Check		OK	00:32
Air Blank		0.000	00:32
Control Test		0.081	00:33
Air Blank		0.000	00:33
Subject Sample	# 1	0.096	00:35
Breath Volume		1.796	00:35
Air Blank		0.000	00:36
Air Blank		0.000	00:37
Subject Sample	# 2	0.094	00:39
Breath Volume		1.781	00:39
Air Blank		0.000	00:39
Control Test		0.079	00:40
Air Blank		0.000	00:40
Diagnostics Check		OK	00:40

Data Download Date/Time: 11/17/2017 13:00

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SCHNEIDER,JOSHUA
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 10/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:22

SUBJECT'S LAST NAME: IMHOLT

DATE/TIME OF TEST: 11/10/2017 04:50:52

Diagnostics Check		OK	04:44
Air Blank		0.000	04:45
Control Test		0.081	04:45
Air Blank		0.000	04:45
Subject Sample	# 1	0.012	04:46
Breath Volume		2.101	04:46
Air Blank		0.000	04:47
Air Blank		0.000	04:48
Subject Sample	# 2	0.010	04:49
Breath Volume		1.671	04:49
Air Blank		0.000	04:49
Control Test		0.080	04:50
Air Blank		0.000	04:50
Diagnostics Check		OK	04:50

Data Download Date/Time: 11/17/2017 13:00

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FICARRA,CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 10/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 19:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:50

SUBJECT'S LAST NAME: WEBSTER

DATE/TIME OF TEST: 11/10/2017 20:19:35

Diagnostics Check		OK	20:12
Air Blank		0.000	20:13
Control Test		0.080	20:13
Air Blank		0.000	20:14
Subject Sample	# 1	0.160	20:15
Breath Volume		2.746	20:15
Air Blank		0.000	20:15
Air Blank		0.000	20:17
Subject Sample	# 2	0.160	20:17
Breath Volume		2.492	20:17
Air Blank		0.000	20:18
Control Test		0.079	20:18
Air Blank		0.000	20:19
Diagnostics Check		OK	20:19

Data Download Date/Time: 11/17/2017 13:00

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: TAVARES,LUCAS
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 10/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:02

SUBJECT'S LAST NAME: MARTINEZ-MASRIEGOS

DATE/TIME OF TEST: 11/11/2017 03:33:32

Diagnostics Check		OK	03:26
Air Blank		0.000	03:26
Control Test		0.080	03:27
Air Blank		0.000	03:27
Subject Sample	# 1	0.147	03:28
Breath Volume		3.066	03:28
Air Blank		0.000	03:29
Air Blank		0.000	03:31
Subject Sample	# 2	0.148	03:31
Breath Volume		3.589	03:31
Air Blank		0.000	03:32
Control Test		0.080	03:32
Air Blank		0.000	03:33
Diagnostics Check		OK	03:33

Data Download Date/Time: 11/17/2017 13:00

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: YACINTHE,WESTLY
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 10/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:47

SUBJECT'S LAST NAME: PARSON

DATE/TIME OF TEST: 11/11/2017 04:18:02

Diagnostics Check		OK	04:11
Air Blank		0.000	04:12
Control Test		0.080	04:12
Air Blank		0.000	04:12
Subject Sample	# 1	0.131	04:13
Breath Volume		2.000	04:13
Air Blank		0.000	04:14
Air Blank		0.000	04:16
Subject Sample	# 2	0.132	04:16
Breath Volume		1.714	04:16
Air Blank		0.000	04:17
Control Test		0.080	04:17
Air Blank		0.000	04:17
Diagnostics Check		OK	04:18

Data Download Date/Time: 11/17/2017 13:00

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CLEARY,GUS
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 10/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:21

SUBJECT'S LAST NAME: JIMENEZ-LOPEZ

DATE/TIME OF TEST: 11/11/2017 23:59:44

Diagnostics Check		OK	23:51
Air Blank		0.000	23:52
Control Test		0.081	23:52
Air Blank		0.000	23:53
Subject Sample	# 1	0.168	23:54
Breath Volume		1.980	23:54
Air Blank		0.000	23:55
Air Blank		0.000	23:57
Subject Sample	# 2	0.163	23:58
Breath Volume		1.777	23:58
Air Blank		0.000	23:58
Control Test		0.079	23:59
Air Blank		0.000	23:59
Diagnostics Check		OK	23:59

Data Download Date/Time: 11/17/2017 13:00

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: HANTON,MELINDA
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 10/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:27

SUBJECT'S LAST NAME: ANAND

DATE/TIME OF TEST: 11/12/2017 00:59:12

Diagnostics Check		OK	00:51
Air Blank		0.000	00:51
Control Test		0.080	00:52
Air Blank		0.000	00:52
Subject Sample	# 1	0.157	00:53
Breath Volume		1.351	00:53
Air Blank		0.000	00:54
Air Blank		0.000	00:56
Subject Sample	# 2	0.154	00:57
Breath Volume		1.324	00:57
Air Blank		0.000	00:58
Control Test		0.079	00:58
Air Blank		0.000	00:59
Diagnostics Check		OK	00:59

Data Download Date/Time: 11/17/2017 13:00

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FICARRA,CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 10/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:21

SUBJECT'S LAST NAME: CROSS

DATE/TIME OF TEST: 11/12/2017 01:54:06

Diagnostics Check		OK	01:46
Air Blank		0.000	01:47
Control Test		0.080	01:47
Air Blank		0.000	01:48
Subject Sample	# 1	0.215	01:49
Breath Volume		1.695	01:49
Air Blank		0.000	01:49
Air Blank		0.000	01:51
Subject Sample	# 2	0.208	01:52
Breath Volume		1.453	01:52
Air Blank		0.000	01:53
Control Test		0.079	01:53
Air Blank		0.000	01:54
Diagnostics Check		OK	01:54

Data Download Date/Time: 11/17/2017 13:00

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LEVEY,SCOTT
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 10/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:46

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 11/12/2017 05:16:42

Diagnostics Check		OK	05:09
Air Blank		0.000	05:10
Control Test		0.080	05:10
Air Blank		0.000	05:11
Subject Sample	# 1	0.060	05:12
Breath Volume		1.414	05:12
Air Blank		0.000	05:12
Air Blank		0.000	05:14
Subject Sample	# 2	0.063	05:15
Breath Volume		2.824	05:15
Air Blank		0.000	05:15
Control Test		0.080	05:16
Air Blank		0.000	05:16
Diagnostics Check		OK	05:16

Data Download Date/Time: 11/17/2017 13:00

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALMER,SAMANTHA,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SALA,ANTHONY
ARRESTING OFFICER AGENCY: Delray Beach PD
LAST AGENCY INSPECTION: 10/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 09:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:50

SUBJECT'S LAST NAME: DALTON

DATE/TIME OF TEST: 11/12/2017 13:56:50

Diagnostics Check		OK	13:50
Air Blank		0.000	13:50
Control Test		0.080	13:51
Air Blank		0.000	13:51
Subject Sample	# 1	0.000	13:52
Breath Volume		4.421	13:52
Air Blank		0.000	13:52
Air Blank		0.000	13:54
Subject Sample	# 2	0.000	13:55
Breath Volume		3.492	13:55
Air Blank		0.000	13:55
Control Test		0.080	13:56
Air Blank		0.000	13:56
Diagnostics Check		OK	13:56

Data Download Date/Time: 11/17/2017 13:00

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALMER,SAMANTHA,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FLEMING,JOHN
ARRESTING OFFICER AGENCY: Juno Beach PD
LAST AGENCY INSPECTION: 10/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 14:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:00

SUBJECT'S LAST NAME: DISE

DATE/TIME OF TEST: 11/12/2017 16:57:46

Diagnostics Check		OK	16:50
Air Blank		0.000	16:51
Control Test		0.080	16:51
Air Blank		0.000	16:52
Subject Sample	# 1	0.263	16:53
Breath Volume		2.597	16:53
Air Blank		0.000	16:53
Air Blank		0.000	16:55
Subject Sample	# 2	0.271	16:56
Breath Volume		2.898	16:56
Air Blank		0.000	16:56
Control Test		0.079	16:57
Air Blank		0.000	16:57
Diagnostics Check		OK	16:57

Data Download Date/Time: 11/17/2017 13:00

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALMER,SAMANTHA,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CHIQUITO RODRIGUEZ,LENDY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 10/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 19:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:55

SUBJECT'S LAST NAME: DANIELS

DATE/TIME OF TEST: 11/12/2017 21:26:51

Diagnostics Check		OK	21:20
Air Blank		0.000	21:20
Control Test		0.081	21:20
Air Blank		0.000	21:21
Subject Sample	# 1	0.232	21:22
Breath Volume		2.070	21:22
Air Blank		0.000	21:23
Air Blank		0.000	21:24
Subject Sample	# 2	0.234	21:25
Breath Volume		1.164	21:25
Air Blank		0.000	21:25
Control Test		0.079	21:26
Air Blank		0.000	21:26
Diagnostics Check		OK	21:26

Data Download Date/Time: 11/17/2017 13:00

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CONNACHER,JOHN
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 10/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:07

SUBJECT'S LAST NAME: LOPEZ VASQUEZ

DATE/TIME OF TEST: 11/12/2017 23:40:36

Diagnostics Check		OK	23:33
Air Blank		0.000	23:33
Control Test		0.080	23:34
Air Blank		0.000	23:34
Subject Sample	# 1	0.167	23:35
Breath Volume		1.285	23:35
Air Blank		0.000	23:36
Air Blank		0.000	23:38
Subject Sample	# 2	0.168	23:38
Breath Volume		1.425	23:38
Air Blank		0.000	23:39
Control Test		0.078	23:39
Air Blank		0.000	23:40
Diagnostics Check		OK	23:40

Data Download Date/Time: 11/17/2017 13:00

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALMER,SAMANTHA,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MOLINARO,TAYLOR
ARRESTING OFFICER AGENCY: Palm Beach PD
LAST AGENCY INSPECTION: 10/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 07:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:20

SUBJECT'S LAST NAME: ORTIZ

DATE/TIME OF TEST: 11/15/2017 09:01:53

Diagnostics Check		OK	08:55
Air Blank		0.000	08:55
Control Test		0.081	08:56
Air Blank		0.000	08:56
Subject Sample	# 1	0.000	08:57
Breath Volume		1.914	08:57
Air Blank		0.000	08:57
Air Blank		0.000	08:59
Subject Sample	# 2	0.000	09:00
Breath Volume		2.460	09:00
Air Blank		0.000	09:00
Control Test		0.081	09:01
Air Blank		0.000	09:01
Diagnostics Check		OK	09:01

Data Download Date/Time: 11/17/2017 13:01

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARI,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: HANTON,MELINDA
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 10/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: EDELMAN

DATE/TIME OF TEST: 11/16/2017 02:06:00

Diagnostics Check		OK	01:59
Air Blank		0.000	01:59
Control Test		0.080	02:00
Air Blank		0.000	02:00
Subject Sample	# 1	0.196	02:01
Breath Volume		2.007	02:01
Air Blank		0.000	02:02
Air Blank		0.000	02:03
Subject Sample	# 2	0.184	02:04
Breath Volume		2.792	02:04
Air Blank		0.000	02:05
Control Test		0.080	02:05
Air Blank		0.000	02:05
Diagnostics Check		OK	02:06

Data Download Date/Time: 11/17/2017 13:01

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARI,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITE,EDWARD
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 10/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:58

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 11/16/2017 03:32:31

Diagnostics Check		OK	03:26
Air Blank		0.000	03:26
Control Test		0.080	03:26
Air Blank		0.000	03:27
Subject Sample	# 1	0.159	03:27
Breath Volume		1.605	03:27
Air Blank		0.000	03:28
Air Blank		0.000	03:30
Subject Sample	# 2	0.154	03:30
Breath Volume		1.335	03:30
Air Blank		0.000	03:31
Control Test		0.080	03:31
Air Blank		0.000	03:32
Diagnostics Check		OK	03:32

Data Download Date/Time: 11/17/2017 13:01

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARI,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BORROWS,ANDREW
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 10/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: CRIPPS

DATE/TIME OF TEST: 11/16/2017 04:49:49

Diagnostics Check		OK	04:43
Air Blank		0.000	04:43
Control Test		0.080	04:44
Air Blank		0.000	04:44
Subject Sample	# 1	0.111	04:45
Breath Volume		1.742	04:45
Air Blank		0.000	04:46
Air Blank		0.000	04:47
Subject Sample	# 2	0.102	04:48
Breath Volume		1.296	04:48
Air Blank		0.000	04:48
Control Test		0.080	04:49
Air Blank		0.000	04:49
Diagnostics Check		OK	04:49

Data Download Date/Time: 11/17/2017 13:01

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SORIANO,REYNALDO
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 10/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: LOBO MORALES

DATE/TIME OF TEST: 11/17/2017 02:58:58

Diagnostics Check		OK	02:51
Air Blank		0.000	02:52
Control Test		0.080	02:52
Air Blank		0.000	02:53
Subject Sample	# 1	0.154	02:54
Breath Volume		2.144	02:54
Air Blank		0.000	02:55
Air Blank		0.000	02:56
Subject Sample	# 2	0.152	02:57
Breath Volume		1.109	02:57
Air Blank		0.000	02:58
Control Test		0.077	02:58
Air Blank		0.000	02:58
Diagnostics Check		OK	02:58

Data Download Date/Time: 11/17/2017 13:01

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FANDREY,CHRISTOPHER
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:05

SUBJECT'S LAST NAME: COGNETTI

DATE/TIME OF TEST: 11/17/2017 23:40:30

Diagnostics Check		OK	23:29
Air Blank		0.000	23:29
Control Test		0.080	23:30
Air Blank		0.000	23:30
Subject Sample	# 1	0.140	23:33
Breath Volume		1.269	23:33
Air Blank		0.000	23:34
Air Blank		0.000	23:36
Subject Sample	# 2	0.153	23:38
Breath Volume		1.183	23:38
Air Blank		0.000	23:39
Control Test		0.078	23:39
Air Blank		0.000	23:40
Diagnostics Check		OK	23:40

Data Download Date/Time: 12/15/2017 15:08

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LEVEY,SCOTT
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: BAKER

DATE/TIME OF TEST: 11/18/2017 01:37:29

Diagnostics Check		OK	01:32	
Air Blank		0.000	01:32	
Control Test		0.079	01:32	
Air Blank		0.000	01:33	
Subject Sample	# 1		01:36	Subject Test Refused
Breath Volume		0.000	01:36	
Air Blank		0.000	01:36	
Control Test		0.080	01:36	
Air Blank		0.000	01:37	
Diagnostics Check		OK	01:37	

Data Download Date/Time: 12/15/2017 15:08

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: GUEVARA,YASEL
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: CALIHAN

DATE/TIME OF TEST: 11/18/2017 02:49:56

Diagnostics Check		OK	02:42
Air Blank		0.000	02:42
Control Test		0.080	02:43
Air Blank		0.000	02:43
Subject Sample	# 1	0.182	02:44
Breath Volume		2.207	02:44
Air Blank		0.000	02:45
Air Blank		0.000	02:47
Subject Sample	# 2	0.166	02:48
Breath Volume		2.027	02:48
Air Blank		0.000	02:48
Control Test		0.079	02:49
Air Blank		0.000	02:49
Diagnostics Check		OK	02:49

Data Download Date/Time: 12/15/2017 15:08

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: DOMINGUEZ,JAYMES
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 05:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:40

SUBJECT'S LAST NAME: ESTIME

DATE/TIME OF TEST: 11/18/2017 07:10:45

Diagnostics Check		OK	07:04
Air Blank		0.000	07:04
Control Test		0.080	07:05
Air Blank		0.000	07:05
Subject Sample	# 1	0.229	07:06
Breath Volume		2.183	07:06
Air Blank		0.000	07:06
Air Blank		0.000	07:08
Subject Sample	# 2	0.237	07:09
Breath Volume		2.542	07:09
Air Blank		0.000	07:09
Control Test		0.079	07:10
Air Blank		0.000	07:10
Diagnostics Check		OK	07:10

Data Download Date/Time: 12/15/2017 15:08

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAIN,JEANETTE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CRUZ,GARDENIA
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 10:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:08

SUBJECT'S LAST NAME: MALDONADO SALES

DATE/TIME OF TEST: 11/18/2017 12:06:44

Diagnostics Check		OK	12:03	
Air Blank		0.000	12:04	
Control Test		0.080	12:04	
Air Blank		0.000	12:04	
Subject Sample	# 1		12:05	Subject Test Refused
Breath Volume		0.000	12:05	
Air Blank		0.000	12:05	
Control Test		0.079	12:06	
Air Blank		0.000	12:06	
Diagnostics Check		OK	12:06	

Data Download Date/Time: 12/15/2017 15:08

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ROTHENBURG,JONATHAN
ARRESTING OFFICER AGENCY: Palm Beach PD
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 18:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:48

SUBJECT'S LAST NAME: MINASIAN

DATE/TIME OF TEST: 11/18/2017 19:28:37

Diagnostics Check		OK	19:15	
Air Blank		0.000	19:15	
Control Test		0.080	19:16	
Air Blank		0.000	19:16	
Subject Sample	# 1	0.249	19:17	No .020 agreement
Breath Volume		1.855	19:17	
Air Blank		0.000	19:18	
Air Blank		0.000	19:20	
Subject Sample	# 2	0.172	19:23	Volume Not Met
Breath Volume		0.488	19:23	
Air Blank		0.000	19:24	
Air Blank		0.000	19:25	
Subject Sample	# 3	0.211	19:26	No .020 agreement
Breath Volume		1.304	19:26	
Air Blank		0.000	19:27	
Control Test		0.077	19:27	
Air Blank		0.000	19:28	
Diagnostics Check		OK	19:28	

Data Download Date/Time: 12/15/2017 15:08

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ROTHENBURG,JONATHAN
ARRESTING OFFICER AGENCY: Palm Beach PD
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 18:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:48

SUBJECT'S LAST NAME: MINASIAN

DATE/TIME OF TEST: 11/18/2017 19:36:34

Diagnostics Check		OK	19:31	
Air Blank		0.000	19:31	
Control Test		0.079	19:32	
Air Blank		0.000	19:32	
Subject Sample	# 1		19:34	Subject Test Refused
Breath Volume		0.156	19:34	
Air Blank		0.000	19:35	
Control Test		0.079	19:35	
Air Blank		0.000	19:36	
Diagnostics Check		OK	19:36	

Data Download Date/Time: 12/15/2017 15:08

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FANDREY,CHRISTOPHER
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:06

SUBJECT'S LAST NAME: LORENTZ

DATE/TIME OF TEST: 11/18/2017 23:42:32

Diagnostics Check		OK	23:35
Air Blank		0.000	23:35
Control Test		0.079	23:35
Air Blank		0.000	23:36
Subject Sample	# 1	0.198	23:37
Breath Volume		1.367	23:37
Air Blank		0.000	23:38
Air Blank		0.000	23:39
Subject Sample	# 2	0.199	23:40
Breath Volume		1.738	23:40
Air Blank		0.000	23:41
Control Test		0.080	23:41
Air Blank		0.000	23:42
Diagnostics Check		OK	23:42

Data Download Date/Time: 12/15/2017 15:08

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SCARTOZZI,PATRICK
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:27

SUBJECT'S LAST NAME: DAVILA HERRERA

DATE/TIME OF TEST: 11/19/2017 01:58:00

Diagnostics Check		OK	01:51
Air Blank		0.000	01:51
Control Test		0.080	01:52
Air Blank		0.000	01:52
Subject Sample	# 1	0.208	01:53
Breath Volume		1.261	01:53
Air Blank		0.000	01:53
Air Blank		0.000	01:55
Subject Sample	# 2	0.217	01:56
Breath Volume		2.339	01:56
Air Blank		0.000	01:57
Control Test		0.080	01:57
Air Blank		0.000	01:57
Diagnostics Check		OK	01:58

Data Download Date/Time: 12/15/2017 15:08

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FANDREY,CHRISTOPHER
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:11

SUBJECT'S LAST NAME: DENECKE

DATE/TIME OF TEST: 11/20/2017 01:48:52

Diagnostics Check		OK	01:37	
Air Blank		0.000	01:37	
Control Test		0.080	01:37	
Air Blank		0.000	01:38	
Subject Sample	# 1	0.168	01:41	Volume Not Met
Breath Volume		0.257	01:41	
Air Blank		0.000	01:42	
Air Blank		0.000	01:43	
Subject Sample	# 2	0.185	01:47	Volume Not Met
Breath Volume		0.113	01:47	
Air Blank		0.000	01:47	
Control Test		0.078	01:48	
Air Blank		0.000	01:48	
Diagnostics Check		OK	01:48	

Data Download Date/Time: 12/15/2017 15:08

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FANDREY,CHRISTOPHER
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:11

SUBJECT'S LAST NAME: DENECKE

DATE/TIME OF TEST: 11/20/2017 02:01:08

Diagnostics Check		OK	01:49	
Air Blank		0.000	01:50	
Control Test		0.078	01:50	
Air Blank		0.000	01:51	
Subject Sample	# 1	0.163	01:54	Volume Not Met
Breath Volume		0.082	01:54	
Air Blank		0.000	01:55	
Air Blank		0.000	01:56	
Subject Sample	# 2		01:59	Subject Test Refused
Breath Volume		0.074	01:59	
Air Blank		0.000	02:00	
Control Test		0.077	02:00	
Air Blank		0.000	02:01	
Diagnostics Check		OK	02:01	

Data Download Date/Time: 12/15/2017 15:08

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAIN,JEANETTE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LAWRENCE,T
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 14:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:39

SUBJECT'S LAST NAME: BAIRD

DATE/TIME OF TEST: 11/20/2017 16:20:04

Diagnostics Check		OK	16:12
Air Blank		0.000	16:13
Control Test		0.079	16:13
Air Blank		0.000	16:14
Subject Sample	# 1	0.000	16:15
Breath Volume		1.328	16:15
Air Blank		0.000	16:15
Air Blank		0.000	16:17
Subject Sample	# 2	0.000	16:18
Breath Volume		1.992	16:18
Air Blank		0.000	16:19
Control Test		0.080	16:19
Air Blank		0.000	16:19
Diagnostics Check		OK	16:20

Data Download Date/Time: 12/15/2017 15:08

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FANDREY,CHRISTOPHER
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: PATAPIS

DATE/TIME OF TEST: 11/20/2017 23:49:02

Diagnostics Check		OK	23:42
Air Blank		0.000	23:43
Control Test		0.079	23:43
Air Blank		0.000	23:44
Subject Sample	# 1	0.000	23:44
Breath Volume		3.468	23:44
Air Blank		0.000	23:45
Air Blank		0.000	23:47
Subject Sample	# 2	0.000	23:47
Breath Volume		3.101	23:47
Air Blank		0.000	23:48
Control Test		0.080	23:48
Air Blank		0.000	23:48
Diagnostics Check		OK	23:49

Data Download Date/Time: 12/15/2017 15:09

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAIN,JEANETTE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FICARRA,CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 19:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:20

SUBJECT'S LAST NAME: KANE

DATE/TIME OF TEST: 11/21/2017 20:52:45

Diagnostics Check		OK	20:46
Air Blank		0.000	20:46
Control Test		0.080	20:46
Air Blank		0.000	20:47
Subject Sample	# 1	0.165	20:48
Breath Volume		1.632	20:48
Air Blank		0.000	20:48
Air Blank		0.000	20:50
Subject Sample	# 2	0.175	20:51
Breath Volume		1.921	20:51
Air Blank		0.000	20:51
Control Test		0.080	20:52
Air Blank		0.000	20:52
Diagnostics Check		OK	20:52

Data Download Date/Time: 12/15/2017 15:09

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAIN,JEANETTE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CHIQUITO-RODRIGUEZ,LENNDY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:26
GENDER: Female

OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: MONOPOLI

DATE/TIME OF TEST: 11/21/2017 22:11:12

Diagnostics Check		OK	22:04
Air Blank		0.000	22:04
Control Test		0.080	22:04
Air Blank		0.000	22:05
Subject Sample	# 1	0.325	22:06
Breath Volume		2.132	22:06
Air Blank		0.000	22:07
Air Blank		0.000	22:08
Subject Sample	# 2	0.314	22:09
Breath Volume		1.156	22:09
Air Blank		0.000	22:10
Control Test		0.078	22:10
Air Blank		0.000	22:11
Diagnostics Check		OK	22:11

Data Download Date/Time: 12/15/2017 15:09

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARI,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SCHAEFER,JOHN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: MANN

DATE/TIME OF TEST: 11/22/2017 00:31:59

Diagnostics Check		OK	00:25
Air Blank		0.000	00:26
Control Test		0.079	00:26
Air Blank		0.000	00:27
Subject Sample	# 1	0.179	00:27
Breath Volume		1.664	00:27
Air Blank		0.000	00:28
Air Blank		0.000	00:29
Subject Sample	# 2	0.173	00:30
Breath Volume		1.515	00:30
Air Blank		0.000	00:31
Control Test		0.079	00:31
Air Blank		0.000	00:31
Diagnostics Check		OK	00:31

Data Download Date/Time: 12/15/2017 15:09

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARI,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: HERNANDEZ,NICOLAS
ARRESTING OFFICER AGENCY: Delray Beach PD
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: BRAVO

DATE/TIME OF TEST: 11/22/2017 04:06:41

Diagnostics Check		OK	04:00
Air Blank		0.000	04:00
Control Test		0.080	04:01
Air Blank		0.000	04:01
Subject Sample	# 1	0.183	04:01
Breath Volume		1.113	04:01
Air Blank		0.000	04:02
Air Blank		0.000	04:04
Subject Sample	# 2	0.183	04:05
Breath Volume		2.343	04:05
Air Blank		0.000	04:05
Control Test		0.079	04:06
Air Blank		0.000	04:06
Diagnostics Check		OK	04:06

Data Download Date/Time: 12/15/2017 15:09

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARI,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: STREIGOLD,DAVID
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: STUART

DATE/TIME OF TEST: 11/24/2017 01:10:38

Diagnostics Check		OK	01:03
Air Blank		0.000	01:04
Control Test		0.080	01:04
Air Blank		0.000	01:05
Subject Sample	# 1	0.181	01:06
Breath Volume		2.207	01:06
Air Blank		0.000	01:06
Air Blank		0.000	01:08
Subject Sample	# 2	0.182	01:08
Breath Volume		2.234	01:08
Air Blank		0.000	01:09
Control Test		0.078	01:10
Air Blank		0.000	01:10
Diagnostics Check		OK	01:10

Data Download Date/Time: 12/15/2017 15:09

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LEVEY,SCOTT
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:07

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 11/24/2017 02:41:57

Diagnostics Check		OK	02:34
Air Blank		0.000	02:34
Control Test		0.080	02:35
Air Blank		0.000	02:35
Subject Sample	# 1	0.187	02:37
Breath Volume		1.980	02:37
Air Blank		0.000	02:37
Air Blank		0.000	02:39
Subject Sample	# 2	0.188	02:40
Breath Volume		1.464	02:40
Air Blank		0.000	02:41
Control Test		0.078	02:41
Air Blank		0.000	02:41
Diagnostics Check		OK	02:41

Data Download Date/Time: 12/15/2017 15:09

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ROTHENBURG,JONATHAN
ARRESTING OFFICER AGENCY: Palm Beach PD
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 19:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:28

SUBJECT'S LAST NAME: NEBEL

DATE/TIME OF TEST: 11/24/2017 21:08:02

Diagnostics Check		OK	20:56	
Air Blank		0.000	20:56	
Control Test		0.081	20:57	
Air Blank		0.000	20:57	
Subject Sample	# 1	0.186	21:00	Volume Not Met
Breath Volume		0.625	21:00	
Air Blank		0.000	21:01	
Air Blank		0.000	21:03	
Subject Sample	# 2	0.191	21:06	Volume Not Met
Breath Volume		0.972	21:06	
Air Blank		0.000	21:07	
Control Test		0.077	21:07	
Air Blank		0.000	21:07	
Diagnostics Check		OK	21:08	

Data Download Date/Time: 12/15/2017 15:09

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ROTHENBURG,JONATHAN
ARRESTING OFFICER AGENCY: Palm Beach PD
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 19:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:28

SUBJECT'S LAST NAME: NEBEL

DATE/TIME OF TEST: 11/24/2017 21:18:47

Diagnostics Check		OK	21:09
Air Blank		0.000	21:10
Control Test		0.078	21:10
Air Blank		0.000	21:11
Subject Sample	# 1	0.202	21:13
Breath Volume		1.882	21:13
Air Blank		0.000	21:14
Air Blank		0.000	21:15
Subject Sample	# 2	0.202	21:17
Breath Volume		2.074	21:17
Air Blank		0.000	21:17
Control Test		0.077	21:18
Air Blank		0.000	21:18
Diagnostics Check		OK	21:18

Data Download Date/Time: 12/15/2017 15:09

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WALTON,THOMAS
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:22

SUBJECT'S LAST NAME: WELCH

DATE/TIME OF TEST: 11/25/2017 04:54:42

Diagnostics Check		OK	04:47
Air Blank		0.000	04:47
Control Test		0.080	04:47
Air Blank		0.000	04:48
Subject Sample	# 1	0.159	04:49
Breath Volume		1.128	04:49
Air Blank		0.000	04:50
Air Blank		0.000	04:52
Subject Sample	# 2	0.163	04:53
Breath Volume		2.023	04:53
Air Blank		0.000	04:53
Control Test		0.079	04:54
Air Blank		0.000	04:54
Diagnostics Check		OK	04:54

Data Download Date/Time: 12/15/2017 15:09

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MC CLUSKEY,RYAN
ARRESTING OFFICER AGENCY: Palm Springs PD
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 19:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:01

SUBJECT'S LAST NAME: PACK

DATE/TIME OF TEST: 11/25/2017 20:34:15

Diagnostics Check		OK	20:27
Air Blank		0.000	20:27
Control Test		0.080	20:28
Air Blank		0.000	20:28
Subject Sample	# 1	0.204	20:29
Breath Volume		1.867	20:29
Air Blank		0.000	20:30
Air Blank		0.000	20:31
Subject Sample	# 2	0.207	20:32
Breath Volume		2.597	20:32
Air Blank		0.000	20:33
Control Test		0.080	20:33
Air Blank		0.000	20:34
Diagnostics Check		OK	20:34

Data Download Date/Time: 12/15/2017 15:09

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CLEARY,GUS
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:39

SUBJECT'S LAST NAME: ORENBUCH

DATE/TIME OF TEST: 11/25/2017 22:11:29

Diagnostics Check		OK	22:04
Air Blank		0.000	22:05
Control Test		0.080	22:05
Air Blank		0.000	22:06
Subject Sample	# 1	0.000	22:07
Breath Volume		3.550	22:07
Air Blank		0.000	22:07
Air Blank		0.000	22:09
Subject Sample	# 2	0.000	22:10
Breath Volume		1.757	22:10
Air Blank		0.000	22:10
Control Test		0.080	22:10
Air Blank		0.000	22:11
Diagnostics Check		OK	22:11

Data Download Date/Time: 12/15/2017 15:09

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SCHNEIDER,JOSHUA
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:46

SUBJECT'S LAST NAME: TEMBRINA

DATE/TIME OF TEST: 11/26/2017 00:17:12

Diagnostics Check		OK	00:09
Air Blank		0.000	00:10
Control Test		0.080	00:10
Air Blank		0.000	00:11
Subject Sample	# 1	0.256	00:12
Breath Volume		2.210	00:12
Air Blank		0.000	00:13
Air Blank		0.000	00:14
Subject Sample	# 2	0.265	00:15
Breath Volume		1.781	00:15
Air Blank		0.000	00:16
Control Test		0.078	00:16
Air Blank		0.000	00:17
Diagnostics Check		OK	00:17

Data Download Date/Time: 12/15/2017 15:09

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LERNER,GORDON
ARRESTING OFFICER AGENCY: Palm Beach Shores
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: TAYLOR-CASTAGNO

DATE/TIME OF TEST: 11/26/2017 03:52:02

Diagnostics Check		OK	03:44
Air Blank		0.000	03:44
Control Test		0.080	03:45
Air Blank		0.000	03:45
Subject Sample	# 1	0.159	03:47
Breath Volume		2.441	03:47
Air Blank		0.000	03:47
Air Blank		0.000	03:49
Subject Sample	# 2	0.155	03:50
Breath Volume		2.257	03:50
Air Blank		0.000	03:51
Control Test		0.079	03:51
Air Blank		0.000	03:51
Diagnostics Check		OK	03:52

Data Download Date/Time: 12/15/2017 15:09

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: PEREZ,ANTONIO
ARRESTING OFFICER AGENCY: Palm Springs PD
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 17:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:42

SUBJECT'S LAST NAME: ABBOTT

DATE/TIME OF TEST: 11/26/2017 19:21:04

Diagnostics Check		OK	19:10
Air Blank		0.000	19:11
Control Test		0.080	19:11
Air Blank		0.000	19:12
Subject Sample	# 1	0.316	19:12
Breath Volume		2.292	19:12
Air Blank		0.000	19:13
Air Blank		0.000	19:15
Subject Sample	# 2	0.289	19:15
Breath Volume		2.054	19:15
Air Blank		0.000	19:16
Air Blank		0.000	19:18
Subject Sample	# 3	0.296	19:19
Breath Volume		1.789	19:19
Air Blank		0.000	19:20
Control Test		0.079	19:20
Air Blank		0.000	19:20
Diagnostics Check		OK	19:21

Data Download Date/Time: 12/15/2017 15:09

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITE,EDWARD
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 18:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:31

SUBJECT'S LAST NAME: WELLINGTON

DATE/TIME OF TEST: 11/26/2017 20:07:44

Diagnostics Check		OK	19:57
Air Blank		0.000	19:57
Control Test		0.079	19:58
Air Blank		0.000	19:58
Subject Sample	# 1	0.231	19:59
Breath Volume		3.085	19:59
Air Blank		0.000	20:00
Air Blank		0.000	20:02
Subject Sample	# 2	0.204	20:02
Breath Volume		1.238	20:02
Air Blank		0.000	20:03
Air Blank		0.000	20:05
Subject Sample	# 3	0.225	20:06
Breath Volume		3.371	20:06
Air Blank		0.000	20:06
Control Test		0.078	20:07
Air Blank		0.000	20:07
Diagnostics Check		OK	20:07

Data Download Date/Time: 12/15/2017 15:09

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAIN,JEANETTE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CHIQUITO-RODRIGUEZ,LENNDY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:00
GENDER: Male

OBSERVATION PERIOD BEGAN: 20:25

SUBJECT'S LAST NAME: LIZANO

DATE/TIME OF TEST: 11/28/2017 20:59:12

Diagnostics Check		OK	20:51
Air Blank		0.000	20:52
Control Test		0.080	20:52
Air Blank		0.000	20:53
Subject Sample	# 1	0.136	20:54
Breath Volume		2.394	20:54
Air Blank		0.000	20:55
Air Blank		0.000	20:57
Subject Sample	# 2	0.123	20:57
Breath Volume		1.371	20:57
Air Blank		0.000	20:58
Control Test		0.080	20:58
Air Blank		0.000	20:59
Diagnostics Check		OK	20:59

Data Download Date/Time: 12/15/2017 15:09

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BORROWS,ADREW
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 18:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:43

SUBJECT'S LAST NAME: SWALM

DATE/TIME OF TEST: 11/30/2017 19:22:06

Diagnostics Check		OK	19:15
Air Blank		0.000	19:15
Control Test		0.081	19:16
Air Blank		0.000	19:16
Subject Sample	# 1	0.327	19:17
Breath Volume		1.917	19:17
Air Blank		0.000	19:18
Air Blank		0.000	19:19
Subject Sample	# 2	0.320	19:20
Breath Volume		1.609	19:20
Air Blank		0.000	19:21
Control Test		0.079	19:21
Air Blank		0.000	19:22
Diagnostics Check		OK	19:22

Data Download Date/Time: 12/15/2017 15:09

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARI,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ANNESE,MORRIS
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:15

SUBJECT'S LAST NAME: GOMEZ

DATE/TIME OF TEST: 11/30/2017 22:49:14

Diagnostics Check		OK	22:42
Air Blank		0.000	22:42
Control Test		0.080	22:42
Air Blank		0.000	22:43
Subject Sample	# 1	0.196	22:44
Breath Volume		2.808	22:44
Air Blank		0.000	22:45
Air Blank		0.000	22:47
Subject Sample	# 2	0.190	22:47
Breath Volume		2.375	22:47
Air Blank		0.000	22:48
Control Test		0.078	22:48
Air Blank		0.000	22:49
Diagnostics Check		OK	22:49

Data Download Date/Time: 12/15/2017 15:09

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KENYAN,M,A
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: KENYAN,MA
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 10/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1841081
DRY GAS EXPIRY DATE: 04/08/2018
VIOLATION CODE: DUI
ARREST TIME: 18:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:30

SUBJECT'S LAST NAME: OURK

DATE/TIME OF TEST: 11/02/2017 18:57:18

Diagnostics Check		OK	18:51
Air Blank		0.000	18:51
Control Test		0.079	18:52
Air Blank		0.000	18:52
Subject Sample	# 1	0.000	18:52
Breath Volume		1.726	18:52
Air Blank		0.000	18:53
Air Blank		0.000	18:55
Subject Sample	# 2	0.000	18:55
Breath Volume		1.292	18:55
Air Blank		0.000	18:56
Control Test		0.078	18:56
Air Blank		0.000	18:57
Diagnostics Check		OK	18:57

Data Download Date/Time: 11/27/2017 11:55

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 10/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1841081
DRY GAS EXPIRY DATE: 04/08/2018
VIOLATION CODE: DUI
ARREST TIME: 21:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:41

SUBJECT'S LAST NAME: REEVES

DATE/TIME OF TEST: 11/03/2017 22:09:02

Diagnostics Check		OK	22:01
Air Blank		0.000	22:01
Control Test		0.079	22:02
Air Blank		0.000	22:02
Subject Sample	# 1	0.064	22:03
Breath Volume		1.191	22:03
Air Blank		0.000	22:04
Air Blank		0.000	22:06
Subject Sample	# 2	0.061	22:07
Breath Volume		1.714	22:07
Air Blank		0.000	22:08
Control Test		0.077	22:08
Air Blank		0.000	22:08
Diagnostics Check		OK	22:09

Data Download Date/Time: 11/27/2017 11:55

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLEMAN,WILLIAM,D
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: WHITTAKER,BRIANNA
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 10/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1841081
DRY GAS EXPIRY DATE: 04/08/2018
VIOLATION CODE: DUI
ARREST TIME: 08:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:42

SUBJECT'S LAST NAME: BURNHAM

DATE/TIME OF TEST: 11/04/2017 09:09:48

Diagnostics Check		OK	09:03
Air Blank		0.000	09:03
Control Test		0.079	09:03
Air Blank		0.000	09:04
Subject Sample	# 1	0.121	09:04
Breath Volume		1.621	09:04
Air Blank		0.000	09:05
Air Blank		0.000	09:07
Subject Sample	# 2	0.115	09:08
Breath Volume		1.386	09:08
Air Blank		0.000	09:08
Control Test		0.079	09:09
Air Blank		0.000	09:09
Diagnostics Check		OK	09:09

Data Download Date/Time: 11/27/2017 11:55

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 10/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1841081
DRY GAS EXPIRY DATE: 04/08/2018
VIOLATION CODE: DUI
ARREST TIME: 22:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:34

SUBJECT'S LAST NAME: PEREZ-SANCHEZ

DATE/TIME OF TEST: 11/05/2017 23:02:33

Diagnostics Check		OK	22:55
Air Blank		0.000	22:56
Control Test		0.079	22:56
Air Blank		0.000	22:57
Subject Sample	# 1	0.324	22:57
Breath Volume		2.449	22:57
Air Blank		0.000	22:58
Air Blank		0.000	23:00
Subject Sample	# 2	0.324	23:00
Breath Volume		3.613	23:00
Air Blank		0.000	23:01
Control Test		0.077	23:01
Air Blank		0.000	23:02
Diagnostics Check		OK	23:02

Data Download Date/Time: 11/27/2017 11:55

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 10/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1841081
DRY GAS EXPIRY DATE: 04/08/2018
VIOLATION CODE: DUI
ARREST TIME: 16:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:55

SUBJECT'S LAST NAME: WALKER

DATE/TIME OF TEST: 11/18/2017 17:21:40

Diagnostics Check		OK	17:15
Air Blank		0.000	17:16
Control Test		0.078	17:16
Air Blank		0.000	17:16
Subject Sample	# 1	0.000	17:17
Breath Volume		1.722	17:17
Air Blank		0.000	17:17
Air Blank		0.000	17:19
Subject Sample	# 2	0.000	17:20
Breath Volume		1.863	17:20
Air Blank		0.000	17:20
Control Test		0.077	17:21
Air Blank		0.000	17:21
Diagnostics Check		OK	17:21

Data Download Date/Time: 11/27/2017 11:55

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAKAI,MICHAEL,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: MAKAI,MICHAEL
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 10/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1841081
DRY GAS EXPIRY DATE: 04/08/2018
VIOLATION CODE: DUI
ARREST TIME: 00:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:32

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 11/23/2017 00:57:27

Diagnostics Check		OK	00:53	
Air Blank		0.000	00:54	
Control Test		0.079	00:54	
Air Blank		0.000	00:55	
Subject Sample	# 1		00:56	Subject Test Refused
Breath Volume		0.000	00:56	
Air Blank		0.000	00:56	
Control Test		0.079	00:56	
Air Blank		0.000	00:57	
Diagnostics Check		OK	00:57	

Data Download Date/Time: 11/27/2017 11:55

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUICK,BENJAMIN,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: QUICK,BENJAMIN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 10/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1841081
DRY GAS EXPIRY DATE: 04/08/2018
VIOLATION CODE: DUI
ARREST TIME: 21:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:55

SUBJECT'S LAST NAME: BENITEZ

DATE/TIME OF TEST: 11/23/2017 22:21:41

Diagnostics Check		OK	22:15
Air Blank		0.000	22:15
Control Test		0.078	22:16
Air Blank		0.000	22:16
Subject Sample	# 1	0.000	22:17
Breath Volume		1.320	22:17
Air Blank		0.000	22:17
Air Blank		0.000	22:19
Subject Sample	# 2	0.000	22:20
Breath Volume		2.078	22:20
Air Blank		0.000	22:20
Control Test		0.080	22:21
Air Blank		0.000	22:21
Diagnostics Check		OK	22:21

Data Download Date/Time: 11/27/2017 11:55

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FITZPATRICK,BRETT,A
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: FITZPATRICK,BRETT
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 10/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 02:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:07

SUBJECT'S LAST NAME: HANKS

DATE/TIME OF TEST: 11/01/2017 03:31:10

Diagnostics Check		OK	03:29	
Air Blank		0.000	03:29	
Control Test		0.081	03:29	
Air Blank		0.000	03:30	
Subject Sample	# 1		03:30	Sequence Aborted
Breath Volume		0.000	03:30	
Air Blank		0.000	03:31	

Data Download Date/Time: 11/27/2017 11:34

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MULLIGAN,WILLIAM,B
BREATH TEST OPERATOR AGENCY: Bradenton PD
ARRESTING OFFICER: MULLIGAN,WILLIAM
ARRESTING OFFICER AGENCY: Bradenton PD
LAST AGENCY INSPECTION: 10/24/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 22:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:37

SUBJECT'S LAST NAME: EATON

DATE/TIME OF TEST: 11/14/2017 00:08:27

Diagnostics Check		OK	00:04	
Air Blank		0.000	00:04	
Control Test		0.081	00:04	
Air Blank		0.000	00:05	
Subject Sample	# 1		00:07	Subject Test Refused
Breath Volume		0.000	00:07	
Air Blank		0.000	00:07	
Control Test		0.081	00:07	
Air Blank		0.000	00:08	
Diagnostics Check		OK	00:08	

Data Download Date/Time: 11/27/2017 11:34

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMPSON,STEVEN,M
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: SIMPSON,STEVEN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 10/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 23:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:32

SUBJECT'S LAST NAME: THEDE

DATE/TIME OF TEST: 11/17/2017 01:10:21

Diagnostics Check		OK	00:58	
Air Blank		0.000	00:59	
Control Test		0.081	00:59	
Air Blank		0.000	01:00	
Subject Sample	# 1		01:03	No Sample Provided
Breath Volume		0.000	01:03	
Air Blank		0.000	01:03	
Air Blank		0.000	01:05	
Subject Sample	# 2		01:08	No Sample Provided
Breath Volume		0.000	01:08	
Air Blank		0.000	01:09	
Control Test		0.082	01:09	
Air Blank		0.000	01:10	
Diagnostics Check		OK	01:10	

Data Download Date/Time: 11/27/2017 11:34

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PETERSON,RONALD,E
BREATH TEST OPERATOR AGENCY: Bradenton PD
ARRESTING OFFICER: AFFOLTER,KRISTIE
ARRESTING OFFICER AGENCY: Bradenton PD
LAST AGENCY INSPECTION: 10/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 19:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:26

SUBJECT'S LAST NAME: VANCE

DATE/TIME OF TEST: 11/23/2017 20:58:33

Diagnostics Check		OK	20:51
Air Blank		0.000	20:52
Control Test		0.081	20:52
Air Blank		0.000	20:52
Subject Sample	# 1	0.253	20:53
Breath Volume		1.277	20:53
Air Blank		0.000	20:54
Air Blank		0.000	20:56
Subject Sample	# 2	0.255	20:56
Breath Volume		3.078	20:56
Air Blank		0.000	20:57
Control Test		0.080	20:57
Air Blank		0.000	20:58
Diagnostics Check		OK	20:58

Data Download Date/Time: 11/27/2017 11:34

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MELENDEZ-BONILLA,RONALD,
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: MELENDEZ-BONILLA,RONALD
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 10/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 20:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: DUNCAN

DATE/TIME OF TEST: 11/24/2017 01:28:51

Diagnostics Check		OK	01:20	
Air Blank		0.000	01:20	
Control Test		0.081	01:20	
Air Blank		0.000	01:21	
Subject Sample	# 1		01:24	No Sample Provided
Breath Volume		0.000	01:24	
Air Blank		0.000	01:25	
Air Blank		0.000	01:27	
Subject Sample	# 2	0.080	01:27	
Breath Volume		3.664	01:27	
Air Blank			01:28	Purge Fail
Air Blank		0.000	01:28	

Data Download Date/Time: 11/27/2017 11:34

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MELENDEZ-BONILLA,RONALD,
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: MELENDEZ-BONILLA,RONALD
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 10/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 20:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: DUNCAN

DATE/TIME OF TEST: 11/24/2017 01:50:45

Diagnostics Check		OK	01:44
Air Blank		0.000	01:44
Control Test		0.081	01:44
Air Blank		0.000	01:45
Subject Sample	# 1	0.076	01:46
Breath Volume		3.632	01:46
Air Blank		0.000	01:46
Air Blank		0.000	01:48
Subject Sample	# 2	0.065	01:49
Breath Volume		1.753	01:49
Air Blank		0.000	01:49
Control Test		0.079	01:50
Air Blank		0.000	01:50
Diagnostics Check		OK	01:50

Data Download Date/Time: 11/27/2017 11:34

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GWODZ,OFC,R
BREATH TEST OPERATOR AGENCY: Bradenton PD
ARRESTING OFFICER: GWODZ,OFC
ARRESTING OFFICER AGENCY: Bradenton PD
LAST AGENCY INSPECTION: 10/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 06:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:07

SUBJECT'S LAST NAME: FRISCH

DATE/TIME OF TEST: 11/25/2017 08:42:22

Diagnostics Check		OK	08:33
Air Blank		0.000	08:34
Control Test		0.081	08:34
Air Blank		0.000	08:35
Subject Sample	# 1	0.213	08:36
Breath Volume		1.664	08:36
Air Blank		0.000	08:37
Air Blank		0.000	08:38
Subject Sample	# 2	0.215	08:40
Breath Volume		2.031	08:40
Air Blank		0.000	08:41
Control Test		0.076	08:41
Air Blank		0.000	08:42
Diagnostics Check		OK	08:42

Data Download Date/Time: 11/27/2017 11:34

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRABAND,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 10/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:47

SUBJECT'S LAST NAME: DUHAIME

DATE/TIME OF TEST: 11/01/2017 02:14:40

Diagnostics Check		OK	02:07
Air Blank		0.000	02:08
Control Test		0.081	02:08
Air Blank		0.000	02:09
Subject Sample	# 1	0.272	02:10
Breath Volume		4.230	02:10
Air Blank		0.000	02:10
Air Blank		0.000	02:12
Subject Sample	# 2	0.278	02:13
Breath Volume		3.476	02:13
Air Blank		0.000	02:13
Control Test		0.079	02:14
Air Blank		0.000	02:14
Diagnostics Check		OK	02:14

Data Download Date/Time: 11/21/2017 14:06

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 10/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:42

SUBJECT'S LAST NAME: KELLY

DATE/TIME OF TEST: 11/02/2017 02:09:35

Diagnostics Check		OK	02:02
Air Blank		0.000	02:03
Control Test		0.081	02:03
Air Blank		0.000	02:04
Subject Sample	# 1	0.202	02:04
Breath Volume		1.125	02:04
Air Blank		0.000	02:05
Air Blank		0.000	02:07
Subject Sample	# 2	0.203	02:07
Breath Volume		1.894	02:07
Air Blank		0.000	02:08
Control Test		0.080	02:08
Air Blank		0.000	02:09
Diagnostics Check		OK	02:09

Data Download Date/Time: 11/21/2017 14:06

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GONZALEZ,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:22

SUBJECT'S LAST NAME: REYNOLDS

DATE/TIME OF TEST: 11/02/2017 22:50:14

Diagnostics Check		OK	22:43
Air Blank		0.000	22:43
Control Test		0.081	22:43
Air Blank		0.000	22:44
Subject Sample	# 1	0.266	22:45
Breath Volume		3.871	22:45
Air Blank		0.000	22:46
Air Blank		0.000	22:48
Subject Sample	# 2	0.263	22:48
Breath Volume		3.425	22:48
Air Blank		0.000	22:49
Control Test		0.079	22:49
Air Blank		0.000	22:50
Diagnostics Check		OK	22:50

Data Download Date/Time: 11/21/2017 14:06

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRIMBLE,TERESA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PORTMAN,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 10/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: YATES

DATE/TIME OF TEST: 11/03/2017 03:42:08

Diagnostics Check		OK	03:35
Air Blank		0.000	03:36
Control Test		0.080	03:36
Air Blank		0.000	03:37
Subject Sample	# 1	0.139	03:37
Breath Volume		2.703	03:37
Air Blank		0.000	03:38
Air Blank		0.000	03:40
Subject Sample	# 2	0.138	03:40
Breath Volume		2.113	03:40
Air Blank		0.000	03:41
Control Test		0.081	03:41
Air Blank		0.000	03:42
Diagnostics Check		OK	03:42

Data Download Date/Time: 11/21/2017 14:06

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SIMPKINS,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:11

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 11/03/2017 23:41:16

Diagnostics Check		OK	23:31
Air Blank		0.000	23:31
Control Test		0.080	23:31
Air Blank		0.000	23:32
Subject Sample	# 1	0.138	23:34
Breath Volume		1.414	23:34
Air Blank		0.000	23:35
Air Blank		0.000	23:37
Subject Sample	# 2	0.135	23:39
Breath Volume		1.507	23:39
Air Blank		0.000	23:40
Control Test		0.079	23:40
Air Blank		0.000	23:41
Diagnostics Check		OK	23:41

Data Download Date/Time: 11/21/2017 14:06

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: COSTA,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:38

SUBJECT'S LAST NAME: RAMOS-BAUTISTA

DATE/TIME OF TEST: 11/04/2017 23:05:27

Diagnostics Check		OK	22:58
Air Blank		0.000	22:59
Control Test		0.081	22:59
Air Blank		0.000	23:00
Subject Sample	# 1	0.160	23:00
Breath Volume		2.468	23:00
Air Blank		0.000	23:01
Air Blank		0.000	23:03
Subject Sample	# 2	0.155	23:03
Breath Volume		1.863	23:03
Air Blank		0.000	23:04
Control Test		0.080	23:04
Air Blank		0.000	23:05
Diagnostics Check		OK	23:05

Data Download Date/Time: 11/21/2017 14:06

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DUNCAN,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:24

SUBJECT'S LAST NAME: CISNEROS-VELASQUEZ

DATE/TIME OF TEST: 11/05/2017 02:51:30

Diagnostics Check		OK	02:44
Air Blank		0.000	02:45
Control Test		0.080	02:45
Air Blank		0.000	02:46
Subject Sample	# 1	0.179	02:46
Breath Volume		1.800	02:46
Air Blank		0.000	02:47
Air Blank		0.000	02:49
Subject Sample	# 2	0.175	02:49
Breath Volume		2.308	02:49
Air Blank		0.000	02:50
Control Test		0.079	02:50
Air Blank		0.000	02:51
Diagnostics Check		OK	02:51

Data Download Date/Time: 11/21/2017 14:06

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PECYLAK,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: LYNCH

DATE/TIME OF TEST: 11/05/2017 04:47:16

Diagnostics Check		OK	04:40
Air Blank		0.000	04:41
Control Test		0.080	04:41
Air Blank		0.000	04:42
Subject Sample	# 1	0.202	04:42
Breath Volume		3.187	04:42
Air Blank		0.000	04:43
Air Blank		0.000	04:45
Subject Sample	# 2	0.199	04:45
Breath Volume		2.851	04:45
Air Blank		0.000	04:46
Control Test		0.079	04:46
Air Blank		0.000	04:47
Diagnostics Check		OK	04:47

Data Download Date/Time: 11/21/2017 14:06

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 10/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: VANCE

DATE/TIME OF TEST: 11/06/2017 02:41:55

Diagnostics Check		OK	02:35
Air Blank		0.000	02:36
Control Test		0.081	02:36
Air Blank		0.000	02:36
Subject Sample	# 1	0.066	02:37
Breath Volume		3.160	02:37
Air Blank		0.000	02:38
Air Blank		0.000	02:39
Subject Sample	# 2	0.065	02:40
Breath Volume		3.015	02:40
Air Blank		0.000	02:40
Control Test		0.080	02:41
Air Blank		0.000	02:41
Diagnostics Check		OK	02:41

Data Download Date/Time: 11/21/2017 14:06

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GONZALEZ,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:23

SUBJECT'S LAST NAME: CHANDLER

DATE/TIME OF TEST: 11/07/2017 00:50:17

Diagnostics Check		OK	00:43
Air Blank		0.000	00:44
Control Test		0.081	00:44
Air Blank		0.000	00:45
Subject Sample	# 1	0.207	00:45
Breath Volume		2.421	00:45
Air Blank		0.000	00:46
Air Blank		0.000	00:48
Subject Sample	# 2	0.195	00:48
Breath Volume		1.917	00:48
Air Blank		0.000	00:49
Control Test		0.080	00:49
Air Blank		0.000	00:50
Diagnostics Check		OK	00:50

Data Download Date/Time: 11/21/2017 14:06

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DEJESUS,V
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:56

SUBJECT'S LAST NAME: GIESEKING

DATE/TIME OF TEST: 11/07/2017 21:26:21

Diagnostics Check		OK	21:16
Air Blank		0.000	21:16
Control Test		0.080	21:17
Air Blank		0.000	21:17
Subject Sample	# 1	0.255	21:18
Breath Volume		4.375	21:18
Air Blank		0.000	21:18
Air Blank		0.000	21:20
Subject Sample	# 2	0.280	21:21
Breath Volume		4.636	21:21
Air Blank		0.000	21:22
Air Blank		0.000	21:23
Subject Sample	# 3	0.277	21:24
Breath Volume		4.820	21:24
Air Blank		0.000	21:25
Control Test		0.079	21:25
Air Blank		0.000	21:26
Diagnostics Check		OK	21:26

Data Download Date/Time: 11/21/2017 14:06

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PECYLAK,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:49

SUBJECT'S LAST NAME: VILLALBA

DATE/TIME OF TEST: 11/09/2017 21:15:42

Diagnostics Check		OK	21:09
Air Blank		0.000	21:09
Control Test		0.081	21:10
Air Blank		0.000	21:10
Subject Sample	# 1	0.202	21:11
Breath Volume		1.214	21:11
Air Blank		0.000	21:11
Air Blank		0.000	21:13
Subject Sample	# 2	0.197	21:14
Breath Volume		1.628	21:14
Air Blank		0.000	21:14
Control Test		0.080	21:15
Air Blank		0.000	21:15
Diagnostics Check		OK	21:15

Data Download Date/Time: 11/21/2017 14:06

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DE JESUS,V
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:53

SUBJECT'S LAST NAME: HERNANDEZ-LUNA

DATE/TIME OF TEST: 11/10/2017 21:21:16

Diagnostics Check		OK	21:14
Air Blank		0.000	21:14
Control Test		0.081	21:14
Air Blank		0.000	21:15
Subject Sample	# 1	0.240	21:16
Breath Volume		2.019	21:16
Air Blank		0.000	21:16
Air Blank		0.000	21:18
Subject Sample	# 2	0.230	21:19
Breath Volume		1.449	21:19
Air Blank		0.000	21:20
Control Test		0.080	21:20
Air Blank		0.000	21:21
Diagnostics Check		OK	21:21

Data Download Date/Time: 11/21/2017 14:06

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HONEYGHON,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: VELASCO

DATE/TIME OF TEST: 11/10/2017 23:44:20

Diagnostics Check		OK	23:37
Air Blank		0.000	23:38
Control Test		0.080	23:38
Air Blank		0.000	23:38
Subject Sample	# 1	0.181	23:39
Breath Volume		1.347	23:39
Air Blank		0.000	23:40
Air Blank		0.000	23:42
Subject Sample	# 2	0.193	23:42
Breath Volume		2.117	23:42
Air Blank		0.000	23:43
Control Test		0.079	23:43
Air Blank		0.000	23:44
Diagnostics Check		OK	23:44

Data Download Date/Time: 11/21/2017 14:06

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WILCHER,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 10/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 11/11/2017 02:06:07

Diagnostics Check		OK	01:52	
Air Blank		0.000	01:52	
Control Test		0.080	01:53	
Air Blank		0.000	01:53	
Subject Sample	# 1		01:56	No Sample Provided
Breath Volume		0.000	01:56	
Air Blank		0.000	01:57	
Air Blank		0.000	01:59	
Subject Sample	# 2	0.202	01:59	
Breath Volume		2.527	01:59	
Air Blank		0.000	02:00	
Air Blank		0.000	02:02	
Subject Sample	# 3	0.200	02:04	
Breath Volume		1.984	02:04	
Air Blank		0.000	02:05	
Control Test		0.079	02:05	
Air Blank		0.000	02:06	
Diagnostics Check		OK	02:06	

Data Download Date/Time: 11/21/2017 14:06

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DE JESUS,V
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:23

SUBJECT'S LAST NAME: RAMSDELL

DATE/TIME OF TEST: 11/11/2017 03:51:00

Diagnostics Check		OK	03:44
Air Blank		0.000	03:45
Control Test		0.080	03:45
Air Blank		0.000	03:45
Subject Sample	# 1	0.064	03:46
Breath Volume		2.910	03:46
Air Blank		0.000	03:47
Air Blank		0.000	03:48
Subject Sample	# 2	0.062	03:49
Breath Volume		2.769	03:49
Air Blank		0.000	03:50
Control Test		0.080	03:50
Air Blank		0.000	03:50
Diagnostics Check		OK	03:51

Data Download Date/Time: 11/21/2017 14:06

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MACIAS,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:28

SUBJECT'S LAST NAME: WRIGHT

DATE/TIME OF TEST: 11/11/2017 04:54:29

Diagnostics Check		OK	04:48
Air Blank		0.000	04:48
Control Test		0.080	04:48
Air Blank		0.000	04:49
Subject Sample	# 1	0.194	04:49
Breath Volume		1.886	04:49
Air Blank		0.000	04:50
Air Blank		0.000	04:52
Subject Sample	# 2	0.188	04:52
Breath Volume		2.234	04:52
Air Blank		0.000	04:53
Control Test		0.081	04:53
Air Blank		0.000	04:54
Diagnostics Check		OK	04:54

Data Download Date/Time: 11/21/2017 14:06

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARR,BARBARA,H
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SINANI,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 06:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:07

SUBJECT'S LAST NAME: AGUILAR TOLEDO

DATE/TIME OF TEST: 11/11/2017 07:36:23

Diagnostics Check		OK	07:27
Air Blank		0.000	07:27
Control Test		0.080	07:28
Air Blank		0.000	07:28
Subject Sample	# 1	0.205	07:29
Breath Volume		1.128	07:29
Air Blank		0.000	07:30
Air Blank		0.000	07:31
Subject Sample	# 2	0.215	07:34
Breath Volume		2.292	07:34
Air Blank		0.000	07:35
Control Test		0.080	07:35
Air Blank		0.000	07:36
Diagnostics Check		OK	07:36

Data Download Date/Time: 11/21/2017 14:06

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRIMBLE,TERESA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MACIAS,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: OGLE JR

DATE/TIME OF TEST: 11/12/2017 03:22:11

Diagnostics Check		OK	03:15
Air Blank		0.000	03:16
Control Test		0.080	03:16
Air Blank		0.000	03:17
Subject Sample	# 1	0.190	03:17
Breath Volume		3.390	03:17
Air Blank		0.000	03:18
Air Blank		0.000	03:20
Subject Sample	# 2	0.191	03:20
Breath Volume		3.500	03:20
Air Blank		0.000	03:21
Control Test		0.080	03:21
Air Blank		0.000	03:22
Diagnostics Check		OK	03:22

Data Download Date/Time: 11/21/2017 14:06

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WILCHER,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 10/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:47

SUBJECT'S LAST NAME: JABER

DATE/TIME OF TEST: 11/12/2017 04:14:05

Diagnostics Check		OK	04:07
Air Blank		0.000	04:08
Control Test		0.081	04:08
Air Blank		0.000	04:09
Subject Sample	# 1	0.105	04:09
Breath Volume		2.101	04:09
Air Blank		0.000	04:10
Air Blank		0.000	04:12
Subject Sample	# 2	0.106	04:12
Breath Volume		1.910	04:12
Air Blank		0.000	04:13
Control Test		0.080	04:13
Air Blank		0.000	04:13
Diagnostics Check		OK	04:14

Data Download Date/Time: 11/21/2017 14:06

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRIMBLE,TERESA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SINANI,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 18:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:07

SUBJECT'S LAST NAME: VANVALKENBURGH

DATE/TIME OF TEST: 11/12/2017 19:33:46

Diagnostics Check		OK	19:27
Air Blank		0.000	19:27
Control Test		0.081	19:28
Air Blank		0.000	19:28
Subject Sample	# 1	0.070	19:29
Breath Volume		2.191	19:29
Air Blank		0.000	19:29
Air Blank		0.000	19:31
Subject Sample	# 2	0.070	19:32
Breath Volume		2.308	19:32
Air Blank		0.000	19:32
Control Test		0.081	19:33
Air Blank		0.000	19:33
Diagnostics Check		OK	19:33

Data Download Date/Time: 11/21/2017 14:06

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRONSON,R
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:04

SUBJECT'S LAST NAME: PERSAUD

DATE/TIME OF TEST: 11/13/2017 02:31:49

Diagnostics Check		OK	02:24
Air Blank		0.000	02:25
Control Test		0.081	02:25
Air Blank		0.000	02:26
Subject Sample	# 1	0.230	02:27
Breath Volume		1.789	02:27
Air Blank		0.000	02:28
Air Blank		0.000	02:29
Subject Sample	# 2	0.232	02:30
Breath Volume		2.460	02:30
Air Blank		0.000	02:30
Control Test		0.080	02:31
Air Blank		0.000	02:31
Diagnostics Check		OK	02:31

Data Download Date/Time: 11/21/2017 14:06

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: COSTA,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:22

SUBJECT'S LAST NAME: WETHERINGTON

DATE/TIME OF TEST: 11/14/2017 01:49:23

Diagnostics Check		OK	01:43
Air Blank		0.000	01:43
Control Test		0.080	01:43
Air Blank		0.000	01:44
Subject Sample	# 1	0.035	01:44
Breath Volume		2.230	01:44
Air Blank		0.000	01:45
Air Blank		0.000	01:47
Subject Sample	# 2	0.034	01:47
Breath Volume		2.574	01:47
Air Blank		0.000	01:48
Control Test		0.081	01:48
Air Blank		0.000	01:49
Diagnostics Check		OK	01:49

Data Download Date/Time: 11/21/2017 14:06

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PECYLAK,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:31

SUBJECT'S LAST NAME: TOWNSEND

DATE/TIME OF TEST: 11/15/2017 02:58:56

Diagnostics Check		OK	02:51
Air Blank		0.000	02:51
Control Test		0.080	02:52
Air Blank		0.000	02:52
Subject Sample	# 1	0.241	02:54
Breath Volume		1.816	02:54
Air Blank		0.000	02:55
Air Blank		0.000	02:56
Subject Sample	# 2	0.240	02:57
Breath Volume		1.382	02:57
Air Blank		0.000	02:57
Control Test		0.080	02:58
Air Blank		0.000	02:58
Diagnostics Check		OK	02:58

Data Download Date/Time: 11/21/2017 14:06

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MACIAS,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: HOANG

DATE/TIME OF TEST: 11/16/2017 03:17:10

Diagnostics Check		OK	03:11
Air Blank		0.000	03:11
Control Test		0.080	03:11
Air Blank		0.000	03:12
Subject Sample	# 1	0.132	03:12
Breath Volume		1.597	03:12
Air Blank		0.000	03:13
Air Blank		0.000	03:15
Subject Sample	# 2	0.126	03:15
Breath Volume		2.593	03:15
Air Blank		0.000	03:16
Control Test		0.080	03:16
Air Blank		0.000	03:17
Diagnostics Check		OK	03:17

Data Download Date/Time: 11/21/2017 14:06

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRIMBLE,TERESA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GONZALEZ,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:22

SUBJECT'S LAST NAME: JEMISON

DATE/TIME OF TEST: 11/16/2017 23:27:59

Diagnostics Check		OK	23:18
Air Blank		0.000	23:19
Control Test		0.080	23:19
Air Blank		0.000	23:20
Subject Sample	# 1	0.164	23:22
Breath Volume		1.140	23:22
Air Blank		0.000	23:23
Air Blank		0.000	23:24
Subject Sample	# 2	0.172	23:26
Breath Volume		1.234	23:26
Air Blank		0.000	23:27
Control Test		0.080	23:27
Air Blank		0.000	23:27
Diagnostics Check		OK	23:27

Data Download Date/Time: 11/21/2017 14:06

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARRELL,Q
ARRESTING OFFICER AGENCY: Plant City PD
LAST AGENCY INSPECTION: 10/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: LUGO

DATE/TIME OF TEST: 11/18/2017 22:33:14

Diagnostics Check		OK	22:26
Air Blank		0.000	22:26
Control Test		0.080	22:27
Air Blank		0.000	22:27
Subject Sample	# 1	0.204	22:28
Breath Volume		1.617	22:28
Air Blank		0.000	22:28
Air Blank		0.000	22:30
Subject Sample	# 2	0.210	22:31
Breath Volume		3.125	22:31
Air Blank		0.000	22:32
Control Test		0.079	22:32
Air Blank		0.000	22:33
Diagnostics Check		OK	22:33

Data Download Date/Time: 11/21/2017 14:06

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ANDERSON,C
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 10/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:17

SUBJECT'S LAST NAME: MYRICK JR

DATE/TIME OF TEST: 11/18/2017 23:54:54

Diagnostics Check		OK	23:37	
Air Blank		0.000	23:38	
Control Test		0.079	23:38	
Air Blank		0.000	23:39	
Subject Sample	# 1	0.169	23:42	Volume Not Met
Breath Volume		0.558	23:42	
Air Blank		0.000	23:42	
Air Blank		0.000	23:44	
Subject Sample	# 2	0.210	23:47	
Breath Volume		1.410	23:47	
Air Blank		0.000	23:48	
Air Blank		0.000	23:50	
Subject Sample	# 3	0.178	23:53	Volume Not Met
Breath Volume		0.875	23:53	
Air Blank		0.000	23:53	
Control Test		0.078	23:54	
Air Blank		0.000	23:54	
Diagnostics Check		OK	23:54	

Data Download Date/Time: 11/21/2017 14:06

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ORNDORFF,W
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: LAYNEZ-RAYMUNDO

DATE/TIME OF TEST: 11/19/2017 03:22:49

Diagnostics Check		OK	03:15
Air Blank		0.000	03:16
Control Test		0.080	03:16
Air Blank		0.000	03:16
Subject Sample	# 1	0.173	03:17
Breath Volume		3.730	03:17
Air Blank		0.000	03:18
Air Blank		0.000	03:20
Subject Sample	# 2	0.172	03:21
Breath Volume		2.175	03:21
Air Blank		0.000	03:21
Control Test		0.079	03:22
Air Blank		0.000	03:22
Diagnostics Check		OK	03:22

Data Download Date/Time: 11/21/2017 14:06

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:57

SUBJECT'S LAST NAME: SAWKA

DATE/TIME OF TEST: 11/19/2017 22:31:44

Diagnostics Check		OK	22:18	
Air Blank		0.000	22:18	
Control Test		0.080	22:18	
Air Blank		0.000	22:19	
Subject Sample	# 1	0.220	22:22	Volume Not Met
Breath Volume		0.839	22:22	
Air Blank		0.000	22:23	
Air Blank		0.000	22:24	
Subject Sample	# 2	0.230	22:25	
Breath Volume		1.183	22:25	
Air Blank		0.000	22:26	
Air Blank		0.000	22:28	
Subject Sample	# 3	0.221	22:29	
Breath Volume		1.359	22:29	
Air Blank		0.000	22:30	
Control Test		0.079	22:31	
Air Blank		0.000	22:31	
Diagnostics Check		OK	22:31	

Data Download Date/Time: 11/21/2017 14:06

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: KRYGIER,B
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 10/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:44

SUBJECT'S LAST NAME: SHEEHAN

DATE/TIME OF TEST: 11/21/2017 02:11:06

Diagnostics Check		OK	02:04
Air Blank		0.000	02:05
Control Test		0.080	02:05
Air Blank		0.000	02:05
Subject Sample	# 1	0.156	02:06
Breath Volume		4.542	02:06
Air Blank		0.000	02:07
Air Blank		0.000	02:08
Subject Sample	# 2	0.151	02:09
Breath Volume		3.851	02:09
Air Blank		0.000	02:10
Control Test		0.079	02:10
Air Blank		0.000	02:11
Diagnostics Check		OK	02:11

Data Download Date/Time: 11/21/2017 14:06

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARR,BARBARA,H
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRENNAN,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 11/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 17:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:02

SUBJECT'S LAST NAME: MUSHRUSH

DATE/TIME OF TEST: 11/30/2017 18:27:20

Diagnostics Check		OK	18:22	
Air Blank		0.000	18:22	
Control Test		0.080	18:22	
Air Blank		0.000	18:23	
Subject Sample	# 1		18:25	Subject Test Refused
Breath Volume		0.000	18:25	
Air Blank		0.000	18:26	
Control Test		0.081	18:26	
Air Blank		0.000	18:27	
Diagnostics Check		OK	18:27	

Data Download Date/Time: 12/19/2017 19:31

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARR,BARBARA,H
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRENNAN,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 11/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 17:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:28

SUBJECT'S LAST NAME: MUSHRUSH

DATE/TIME OF TEST: 11/30/2017 18:55:28

Diagnostics Check		OK	18:49
Air Blank		0.000	18:49
Control Test		0.080	18:49
Air Blank		0.000	18:50
Subject Sample	# 1	0.320	18:50
Breath Volume		2.726	18:50
Air Blank		0.000	18:51
Air Blank		0.000	18:53
Subject Sample	# 2	0.317	18:53
Breath Volume		2.339	18:53
Air Blank		0.000	18:54
Control Test		0.079	18:54
Air Blank		0.000	18:55
Diagnostics Check		OK	18:55

Data Download Date/Time: 12/19/2017 19:31

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRANNIS,RONALD,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: NONE,NONE
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 11/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: ADMIN
ARREST TIME: 23:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: PETERSON

DATE/TIME OF TEST: 11/30/2017 23:40:16

Diagnostics Check		OK	23:35	
Air Blank		0.000	23:35	
Control Test		0.080	23:36	
Air Blank		0.000	23:36	
Subject Sample	# 1	0.234	23:37	
Breath Volume		1.410	23:37	
Air Blank		0.000	23:37	
Air Blank		0.000	23:39	
Subject Sample	# 2		23:39	Sequence Aborted
Breath Volume		0.000	23:39	
Air Blank		0.000	23:40	

Data Download Date/Time: 12/19/2017 19:31

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GADSON,LAUREN,J
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: JUREC,BEN
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 10/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09117080A1
DRY GAS EXPIRY DATE: 06/05/2019
VIOLATION CODE: DUI
ARREST TIME: 15:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:15

SUBJECT'S LAST NAME: DIAZ HERNANDEZ

DATE/TIME OF TEST: 11/01/2017 16:39:43

Diagnostics Check		OK	16:36	
Air Blank		0.000	16:36	
Control Test		0.080	16:36	
Air Blank		0.000	16:37	
Subject Sample	# 1		16:39	RFI Detect
Breath Volume		0.000	16:39	
Air Blank			16:39	RFI Detect
Air Blank			16:39	RFI Detect

Data Download Date/Time: 11/29/2017 22:55

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GADSON,LAUREN,J
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: JUREC,BEN
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 10/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09117080A1
DRY GAS EXPIRY DATE: 06/05/2019
VIOLATION CODE: DUI
ARREST TIME: 15:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:15

SUBJECT'S LAST NAME: DIAZ HERNANDEZ

DATE/TIME OF TEST: 11/01/2017 17:00:44

Diagnostics Check		OK	16:52
Air Blank		0.000	16:52
Control Test		0.079	16:53
Air Blank		0.000	16:53
Subject Sample	# 1	0.153	16:55
Breath Volume		1.824	16:55
Air Blank		0.000	16:56
Air Blank		0.000	16:57
Subject Sample	# 2	0.162	16:59
Breath Volume		1.218	16:59
Air Blank		0.000	16:59
Control Test		0.076	17:00
Air Blank		0.000	17:00
Diagnostics Check		OK	17:00

Data Download Date/Time: 11/29/2017 22:55

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRIEMER,AMANDA,M
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: KULESKI,JONATHAN
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 10/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09117080A1
DRY GAS EXPIRY DATE: 06/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: MCCLUNG

DATE/TIME OF TEST: 11/02/2017 01:18:17

Diagnostics Check		OK	01:15	
Air Blank		0.000	01:15	
Control Test		0.080	01:16	
Air Blank		0.000	01:16	
Subject Sample	# 1		01:16	Subject Test Refused
Breath Volume		0.000	01:16	
Air Blank		0.000	01:17	
Control Test		0.079	01:17	
Air Blank		0.000	01:18	
Diagnostics Check		OK	01:18	

Data Download Date/Time: 11/29/2017 22:55

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GROTH,SHANE,M
BREATH TEST OPERATOR AGENCY: Bunnell PD
ARRESTING OFFICER: GROTH,SHANE
ARRESTING OFFICER AGENCY: Bunnell PD
LAST AGENCY INSPECTION: 10/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09117080A1
DRY GAS EXPIRY DATE: 06/05/2019
VIOLATION CODE: DUI
ARREST TIME: 16:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:15

SUBJECT'S LAST NAME: ZACH

DATE/TIME OF TEST: 11/03/2017 17:42:35

Diagnostics Check		OK	17:36
Air Blank		0.000	17:36
Control Test		0.080	17:37
Air Blank		0.000	17:37
Subject Sample	# 1	0.027	17:38
Breath Volume		1.902	17:38
Air Blank		0.000	17:38
Air Blank		0.000	17:40
Subject Sample	# 2	0.026	17:41
Breath Volume		1.285	17:41
Air Blank		0.000	17:41
Control Test		0.080	17:41
Air Blank		0.000	17:42
Diagnostics Check		OK	17:42

Data Download Date/Time: 11/29/2017 22:55

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NICOLAS,MITA,
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: GADDIE,KYLE
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 10/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09117080A1
DRY GAS EXPIRY DATE: 06/05/2019
VIOLATION CODE: DUI
ARREST TIME: 17:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:35

SUBJECT'S LAST NAME: GETMAN

DATE/TIME OF TEST: 11/03/2017 19:01:45

Diagnostics Check		OK	18:55
Air Blank		0.000	18:55
Control Test		0.079	18:56
Air Blank		0.000	18:56
Subject Sample	# 1	0.221	18:57
Breath Volume		1.632	18:57
Air Blank		0.000	18:57
Air Blank		0.000	18:59
Subject Sample	# 2	0.219	19:00
Breath Volume		1.382	19:00
Air Blank		0.000	19:00
Control Test		0.079	19:01
Air Blank		0.000	19:01
Diagnostics Check		OK	19:01

Data Download Date/Time: 11/29/2017 22:55

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEWIS,CHRISTOPHER,B
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: COZZONE,GAETANO
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 10/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09117080A1
DRY GAS EXPIRY DATE: 06/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: BLACK

DATE/TIME OF TEST: 11/04/2017 00:13:11

Diagnostics Check		OK	00:06
Air Blank		0.000	00:06
Control Test		0.080	00:06
Air Blank		0.000	00:07
Subject Sample	# 1	0.000	00:08
Breath Volume		1.878	00:08
Air Blank		0.000	00:09
Air Blank		0.000	00:11
Subject Sample	# 2	0.000	00:11
Breath Volume		1.328	00:11
Air Blank		0.000	00:12
Control Test		0.077	00:12
Air Blank		0.000	00:13
Diagnostics Check		OK	00:13

Data Download Date/Time: 11/29/2017 22:55

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHEWNING,AUSTIN,M
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: PETKOVSEK,RICHARD
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 10/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09117080A1
DRY GAS EXPIRY DATE: 06/05/2019
VIOLATION CODE: DUI
ARREST TIME: 05:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:47

SUBJECT'S LAST NAME: G656812634280

DATE/TIME OF TEST: 11/04/2017 06:37:08

Diagnostics Check		OK	06:30
Air Blank		0.000	06:31
Control Test		0.080	06:31
Air Blank		0.000	06:32
Subject Sample	# 1	0.223	06:32
Breath Volume		3.175	06:32
Air Blank		0.000	06:33
Air Blank		0.000	06:34
Subject Sample	# 2	0.224	06:35
Breath Volume		2.734	06:35
Air Blank		0.000	06:36
Control Test		0.078	06:36
Air Blank		0.000	06:37
Diagnostics Check		OK	06:37

Data Download Date/Time: 11/29/2017 22:55

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 10/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09117080A1
DRY GAS EXPIRY DATE: 06/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: OFFERMAN

DATE/TIME OF TEST: 11/04/2017 22:31:31

Diagnostics Check		OK	22:25
Air Blank		0.000	22:25
Control Test		0.080	22:25
Air Blank		0.000	22:26
Subject Sample	# 1	0.137	22:26
Breath Volume		1.253	22:26
Air Blank		0.000	22:27
Air Blank		0.000	22:29
Subject Sample	# 2	0.142	22:29
Breath Volume		1.664	22:29
Air Blank		0.000	22:30
Control Test		0.079	22:30
Air Blank		0.000	22:31
Diagnostics Check		OK	22:31

Data Download Date/Time: 11/29/2017 22:55

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRIEMER,AMANDA,M
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: PEDERSEN,ERIK
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 10/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09117080A1
DRY GAS EXPIRY DATE: 06/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: VALLEY

DATE/TIME OF TEST: 11/12/2017 02:39:54

Diagnostics Check		OK	02:33
Air Blank		0.000	02:33
Control Test		0.080	02:34
Air Blank		0.000	02:34
Subject Sample	# 1	0.205	02:35
Breath Volume		1.285	02:35
Air Blank		0.000	02:35
Air Blank		0.000	02:37
Subject Sample	# 2	0.211	02:38
Breath Volume		1.113	02:38
Air Blank		0.000	02:38
Control Test		0.080	02:39
Air Blank		0.000	02:39
Diagnostics Check		OK	02:39

Data Download Date/Time: 11/29/2017 22:55

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NICOLAS,MITA,
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: COCHRAN,WILLIAM
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 10/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09117080A1
DRY GAS EXPIRY DATE: 06/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:15

SUBJECT'S LAST NAME: JORDAN

DATE/TIME OF TEST: 11/14/2017 20:42:14

Diagnostics Check		OK	20:35
Air Blank		0.000	20:36
Control Test		0.079	20:36
Air Blank		0.000	20:37
Subject Sample	# 1	0.197	20:37
Breath Volume		1.609	20:37
Air Blank		0.000	20:38
Air Blank		0.000	20:40
Subject Sample	# 2	0.212	20:40
Breath Volume		1.683	20:40
Air Blank		0.000	20:41
Control Test		0.079	20:41
Air Blank		0.000	20:42
Diagnostics Check		OK	20:42

Data Download Date/Time: 11/29/2017 22:55

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VINCI,ROSANNA,L
BREATH TEST OPERATOR AGENCY: Flagler Beach PD
ARRESTING OFFICER: VINCI,ROSANNA
ARRESTING OFFICER AGENCY: Flagler Beach PD
LAST AGENCY INSPECTION: 10/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09117080A1
DRY GAS EXPIRY DATE: 06/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:18

SUBJECT'S LAST NAME: LONGHWAY

DATE/TIME OF TEST: 11/18/2017 03:41:12

Diagnostics Check		OK	03:38	
Air Blank		0.000	03:38	
Control Test		0.080	03:39	
Air Blank		0.000	03:39	
Subject Sample	# 1		03:39	Subject Test Refused
Breath Volume		0.000	03:39	
Air Blank		0.000	03:40	
Control Test		0.080	03:40	
Air Blank		0.000	03:41	
Diagnostics Check		OK	03:41	

Data Download Date/Time: 11/29/2017 22:55

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PREVATT,JENNIFER,M
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: PREVATT,JENNIFER
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 10/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09117080A1
DRY GAS EXPIRY DATE: 06/05/2019
VIOLATION CODE: DUI
ARREST TIME: 17:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:35

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 11/19/2017 18:05:00

Diagnostics Check		OK	17:58
Air Blank		0.000	17:58
Control Test		0.080	17:58
Air Blank		0.000	17:59
Subject Sample	# 1	0.218	18:00
Breath Volume		2.015	18:00
Air Blank		0.000	18:00
Air Blank		0.000	18:02
Subject Sample	# 2	0.211	18:03
Breath Volume		1.945	18:03
Air Blank		0.000	18:04
Control Test		0.078	18:04
Air Blank		0.000	18:04
Diagnostics Check		OK	18:05

Data Download Date/Time: 11/29/2017 22:55

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEWIS,CHRISTOPHER,B
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: BELL,S
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 10/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09117080A1
DRY GAS EXPIRY DATE: 06/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:19

SUBJECT'S LAST NAME: WALTERS

DATE/TIME OF TEST: 11/23/2017 23:42:57

Diagnostics Check		OK	23:39	
Air Blank		0.000	23:40	
Control Test		0.080	23:40	
Air Blank		0.000	23:40	
Subject Sample	# 1		23:41	Subject Test Refused
Breath Volume		0.000	23:41	
Air Blank		0.000	23:42	
Control Test		0.079	23:42	
Air Blank		0.000	23:42	
Diagnostics Check		OK	23:42	

Data Download Date/Time: 11/29/2017 22:55

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAYLOR,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: GREEN,SETH
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 10/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09117080A1
DRY GAS EXPIRY DATE: 06/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:11

SUBJECT'S LAST NAME: BOGGS

DATE/TIME OF TEST: 11/24/2017 22:35:54

Diagnostics Check		OK	22:32	
Air Blank		0.000	22:33	
Control Test		0.079	22:33	
Air Blank		0.000	22:34	
Subject Sample	# 1		22:34	Subject Test Refused
Breath Volume		0.000	22:34	
Air Blank		0.000	22:34	
Control Test		0.079	22:35	
Air Blank		0.000	22:35	
Diagnostics Check		OK	22:35	

Data Download Date/Time: 11/29/2017 22:55

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEAVER,D,
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: WEAVER,D
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 10/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09117080A1
DRY GAS EXPIRY DATE: 06/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:51

SUBJECT'S LAST NAME: NOBREGA

DATE/TIME OF TEST: 11/26/2017 01:21:51

Diagnostics Check		OK	01:11
Air Blank		0.000	01:12
Control Test		0.080	01:12
Air Blank		0.000	01:13
Subject Sample	# 1	0.217	01:15
Breath Volume		1.312	01:15
Air Blank		0.000	01:16
Air Blank		0.000	01:18
Subject Sample	# 2	0.198	01:20
Breath Volume		1.875	01:20
Air Blank		0.000	01:20
Control Test		0.080	01:21
Air Blank		0.000	01:21
Diagnostics Check		OK	01:21

Data Download Date/Time: 11/29/2017 22:55

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YELVINGTON,AUSTIN,R
BREATH TEST OPERATOR AGENCY: Flagler Beach PD
ARRESTING OFFICER: OLIVA,JOSEPH
ARRESTING OFFICER AGENCY: Flagler Beach PD
LAST AGENCY INSPECTION: 10/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09117080A1
DRY GAS EXPIRY DATE: 06/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:23

SUBJECT'S LAST NAME: MORGAN GRAEFE

DATE/TIME OF TEST: 11/28/2017 04:51:10

Diagnostics Check		OK	04:44
Air Blank		0.000	04:45
Control Test		0.080	04:45
Air Blank		0.000	04:46
Subject Sample	# 1	0.114	04:46
Breath Volume		2.234	04:46
Air Blank		0.000	04:47
Air Blank		0.000	04:49
Subject Sample	# 2	0.113	04:49
Breath Volume		1.593	04:49
Air Blank		0.000	04:50
Control Test		0.080	04:50
Air Blank		0.000	04:51
Diagnostics Check		OK	04:51

Data Download Date/Time: 11/29/2017 22:55

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GADSON,LAUREN,J
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: GREEN,SETH
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 10/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09117080A1
DRY GAS EXPIRY DATE: 06/05/2019
VIOLATION CODE: DUI
ARREST TIME: 19:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:20

SUBJECT'S LAST NAME: LICARI

DATE/TIME OF TEST: 11/29/2017 20:56:06

Diagnostics Check		OK	20:41	
Air Blank		0.000	20:41	
Control Test		0.081	20:41	
Air Blank		0.000	20:42	
Subject Sample	# 1	0.178	20:44	
Breath Volume		1.652	20:44	
Air Blank		0.000	20:46	
Air Blank		0.000	20:47	
Subject Sample	# 2	0.140	20:50	Volume Not Met
Breath Volume		0.800	20:50	
Air Blank		0.000	20:51	
Air Blank		0.000	20:53	
Subject Sample	# 3	0.182	20:54	
Breath Volume		1.777	20:54	
Air Blank		0.000	20:55	
Control Test		0.077	20:55	
Air Blank		0.000	20:56	
Diagnostics Check		OK	20:56	

Data Download Date/Time: 11/29/2017 22:55

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GADSON,LAUREN,J
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: WOODCOCK,MARISSA
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 11/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09117080A1
DRY GAS EXPIRY DATE: 06/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: MILLMAN

DATE/TIME OF TEST: 11/30/2017 03:07:28

Diagnostics Check		OK	03:00
Air Blank		0.000	03:00
Control Test		0.080	03:01
Air Blank		0.000	03:01
Subject Sample	# 1	0.096	03:02
Breath Volume		1.664	03:02
Air Blank		0.000	03:03
Air Blank		0.000	03:05
Subject Sample	# 2	0.088	03:05
Breath Volume		1.492	03:05
Air Blank		0.000	03:06
Control Test		0.079	03:06
Air Blank		0.000	03:07
Diagnostics Check		OK	03:07

Data Download Date/Time: 12/29/2017 16:25

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEEN,ALYIA,S
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: GENDEN,ERIC
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 10/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A1
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 18:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:20

SUBJECT'S LAST NAME: GREGORY

DATE/TIME OF TEST: 11/01/2017 19:23:19

Diagnostics Check		OK	19:20	
Air Blank		0.000	19:20	
Control Test		0.078	19:21	
Air Blank		0.000	19:21	
Subject Sample	# 1		19:21	Subject Test Refused
Breath Volume		0.000	19:21	
Air Blank		0.000	19:22	
Control Test		0.078	19:22	
Air Blank		0.000	19:23	
Diagnostics Check		OK	19:23	

Data Download Date/Time: 11/17/2017 11:06

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FONG,KENNETH,
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: TOWNSLEY,JAMES
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 10/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A1
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: ROBERGE

DATE/TIME OF TEST: 11/05/2017 01:29:12

Diagnostics Check		OK	01:20
Air Blank		0.000	01:21
Control Test		0.078	01:21
Air Blank		0.000	01:22
Subject Sample	# 1	0.097	01:22
Breath Volume		1.449	01:22
Air Blank		0.000	01:23
Air Blank		0.000	01:25
Subject Sample	# 2	0.099	01:27
Breath Volume		1.398	01:27
Air Blank		0.000	01:28
Control Test		0.079	01:28
Air Blank		0.000	01:29
Diagnostics Check		OK	01:29

Data Download Date/Time: 11/17/2017 11:06

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,BRYAN,C
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: GENDEN,ERIC
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 10/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A1
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:05

SUBJECT'S LAST NAME: HANSEN

DATE/TIME OF TEST: 11/07/2017 20:42:03

Diagnostics Check		OK	20:35
Air Blank		0.000	20:36
Control Test		0.077	20:36
Air Blank		0.000	20:37
Subject Sample	# 1	0.000	20:37
Breath Volume		1.429	20:37
Air Blank		0.000	20:38
Air Blank		0.000	20:40
Subject Sample	# 2	0.000	20:40
Breath Volume		1.789	20:40
Air Blank		0.000	20:41
Control Test		0.078	20:41
Air Blank		0.000	20:41
Diagnostics Check		OK	20:42

Data Download Date/Time: 11/17/2017 11:06

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAFALKO,TRAVIS,K
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: BISSOON,STEPHEN
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 10/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A1
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: ALTMAN

DATE/TIME OF TEST: 11/10/2017 02:14:08

Diagnostics Check		OK	02:07
Air Blank		0.000	02:08
Control Test		0.078	02:08
Air Blank		0.000	02:09
Subject Sample	# 1	0.161	02:09
Breath Volume		3.042	02:09
Air Blank		0.000	02:10
Air Blank		0.000	02:12
Subject Sample	# 2	0.155	02:12
Breath Volume		2.617	02:12
Air Blank		0.000	02:13
Control Test		0.079	02:13
Air Blank		0.000	02:14
Diagnostics Check		OK	02:14

Data Download Date/Time: 11/17/2017 11:06

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAFALKO,TRAVIS,K
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: BURNETTE,ASHLEY
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 10/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A1
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:14

SUBJECT'S LAST NAME: MC ENTEE

DATE/TIME OF TEST: 11/10/2017 02:38:45

Diagnostics Check		OK	02:35	
Air Blank		0.000	02:36	
Control Test		0.077	02:36	
Air Blank		0.000	02:37	
Subject Sample	# 1		02:37	Subject Test Refused
Breath Volume		0.000	02:37	
Air Blank		0.000	02:37	
Control Test		0.078	02:38	
Air Blank		0.000	02:38	
Diagnostics Check		OK	02:38	

Data Download Date/Time: 11/17/2017 11:06

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAFALKO,TRAVIS,K
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: KIRK,ANDRAE
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 10/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A1
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: LYDON

DATE/TIME OF TEST: 11/11/2017 00:03:36

Diagnostics Check		OK	23:56
Air Blank		0.000	23:56
Control Test		0.078	23:57
Air Blank		0.000	23:57
Subject Sample	# 2	0.162	23:58
Breath Volume		2.574	23:58
Air Blank		0.000	23:59
Air Blank		0.000	00:01
Subject Sample	# 1	0.169	00:01
Breath Volume		2.652	00:01
Air Blank		0.000	00:02
Control Test		0.077	00:03
Air Blank		0.000	00:03
Diagnostics Check		OK	00:03

Data Download Date/Time: 11/17/2017 11:06

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BISSOON,STEPHEN,R
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: ALVAREZ,AMANDA
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 10/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A1
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:56

SUBJECT'S LAST NAME: GALLON

DATE/TIME OF TEST: 11/11/2017 02:34:44

Diagnostics Check		OK	02:27
Air Blank		0.000	02:27
Control Test		0.078	02:27
Air Blank		0.000	02:28
Subject Sample	# 1	0.183	02:29
Breath Volume		2.039	02:29
Air Blank		0.000	02:30
Air Blank		0.000	02:31
Subject Sample	# 2	0.192	02:32
Breath Volume		2.402	02:32
Air Blank		0.000	02:33
Control Test		0.076	02:34
Air Blank		0.000	02:34
Diagnostics Check		OK	02:34

Data Download Date/Time: 11/17/2017 11:06

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MURPHY,BRITTANIE,J
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: SAAVEDRA,ALONSO
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 10/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A1
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: LIUZZO

DATE/TIME OF TEST: 11/11/2017 03:15:39

Diagnostics Check		OK	03:12	
Air Blank		0.000	03:13	
Control Test		0.077	03:13	
Air Blank		0.000	03:13	
Subject Sample	# 1		03:14	Subject Test Refused
Breath Volume		0.000	03:14	
Air Blank		0.000	03:14	
Control Test		0.078	03:15	
Air Blank		0.000	03:15	
Diagnostics Check		OK	03:15	

Data Download Date/Time: 11/17/2017 11:06

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEEN,ALYIA,S
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: KIRK,ANDRAE
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 10/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A1
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: COLANGELO

DATE/TIME OF TEST: 11/12/2017 01:34:41

Diagnostics Check		OK	01:28
Air Blank		0.000	01:28
Control Test		0.078	01:28
Air Blank		0.000	01:29
Subject Sample	# 1	0.049	01:30
Breath Volume		2.394	01:30
Air Blank		0.000	01:30
Air Blank		0.000	01:32
Subject Sample	# 2	0.046	01:33
Breath Volume		3.648	01:33
Air Blank		0.000	01:33
Control Test		0.075	01:34
Air Blank		0.000	01:34
Diagnostics Check		OK	01:34

Data Download Date/Time: 11/17/2017 11:07

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITE,RYAN,P
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: GENDEN,ERIC
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 10/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A1
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:28

SUBJECT'S LAST NAME: PROCHETTE

DATE/TIME OF TEST: 11/14/2017 22:59:19

Diagnostics Check		OK	22:55	
Air Blank		0.000	22:55	
Control Test		0.078	22:56	
Air Blank		0.000	22:56	
Subject Sample	# 1		22:57	Subject Test Refused
Breath Volume		0.000	22:57	
Air Blank		0.000	22:58	
Control Test		0.079	22:58	
Air Blank		0.000	22:59	
Diagnostics Check		OK	22:59	

Data Download Date/Time: 11/17/2017 11:07

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GENDEN,ERIC,B
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: CROWE,DOUGLAS
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A1
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:26

SUBJECT'S LAST NAME: ROTERMUND

DATE/TIME OF TEST: 11/17/2017 23:54:31

Diagnostics Check		OK	23:48
Air Blank		0.000	23:48
Control Test		0.077	23:48
Air Blank		0.000	23:49
Subject Sample	# 1	0.118	23:49
Breath Volume		1.386	23:49
Air Blank		0.000	23:50
Air Blank		0.000	23:52
Subject Sample	# 2	0.128	23:52
Breath Volume		1.953	23:52
Air Blank		0.000	23:53
Control Test		0.078	23:53
Air Blank		0.000	23:54
Diagnostics Check		OK	23:54

Data Download Date/Time: 12/20/2017 11:34

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITE,RYAN,P
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: GENDEN,ERIC
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 01316080A1
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: TEIXEIRA BARROS

DATE/TIME OF TEST: 11/19/2017 02:51:04

Diagnostics Check		OK	02:37	
Air Blank		0.000	02:38	
Control Test		0.078	02:38	
Air Blank		0.000	02:39	
Subject Sample	# 1		02:42	No Sample Provided
Breath Volume		0.000	02:42	
Air Blank		0.000	02:42	
Air Blank		0.000	02:44	
Subject Sample	# 2	0.069	02:46	
Breath Volume		1.648	02:46	
Air Blank		0.000	02:47	
Air Blank		0.000	02:48	
Subject Sample	# 3	0.063	02:49	
Breath Volume		1.406	02:49	
Air Blank		0.000	02:50	
Control Test		0.078	02:50	
Air Blank		0.000	02:50	
Diagnostics Check		OK	02:51	

Data Download Date/Time: 12/20/2017 11:34

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREENBERG,AMIE,H
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: ZIADIE,ANDREW
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A1
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 14:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:13

SUBJECT'S LAST NAME: SERABIAN

DATE/TIME OF TEST: 11/21/2017 15:50:15

Diagnostics Check		OK	15:43
Air Blank		0.000	15:43
Control Test		0.077	15:44
Air Blank		0.000	15:44
Subject Sample	# 1	0.273	15:45
Breath Volume		2.437	15:45
Air Blank		0.000	15:46
Air Blank		0.000	15:47
Subject Sample	# 2	0.270	15:48
Breath Volume		2.273	15:48
Air Blank		0.000	15:49
Control Test		0.077	15:49
Air Blank		0.000	15:50
Diagnostics Check		OK	15:50

Data Download Date/Time: 12/20/2017 11:34

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEEN,ALYIA,S
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: CASTILLO,JAVIER
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A1
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: FRADLEY

DATE/TIME OF TEST: 11/24/2017 20:28:55

Diagnostics Check		OK	20:22	
Air Blank		0.000	20:22	
Control Test		0.078	20:22	
Air Blank		0.000	20:23	
Subject Sample	# 1	0.225	20:24	
Breath Volume		3.406	20:24	
Air Blank		0.000	20:24	
Air Blank		0.000	20:26	
Subject Sample	# 2	0.218	20:27	
Breath Volume		3.667	20:27	
Air Blank		0.000	20:28	
Control Test		0.074	20:28	Control Outside Tolerance
Air Blank		0.000	20:28	

Data Download Date/Time: 12/20/2017 11:34

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REISSI,DANIEL,J
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: BURNETTE,A
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A1
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 11/24/2017 23:58:46

Diagnostics Check		OK	23:55	
Air Blank		0.000	23:56	
Control Test		0.079	23:56	
Air Blank		0.000	23:57	
Subject Sample	# 1		23:57	Subject Test Refused
Breath Volume		0.000	23:57	
Air Blank		0.000	23:57	
Control Test		0.079	23:58	
Air Blank		0.000	23:58	
Diagnostics Check		OK	23:58	

Data Download Date/Time: 12/20/2017 11:34

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REISSI,DANIEL,J
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: SAAVEDRA,A
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A1
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 18:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:50

SUBJECT'S LAST NAME: VAYNERIS

DATE/TIME OF TEST: 11/25/2017 19:22:17

Diagnostics Check		OK	19:14
Air Blank		0.000	19:14
Control Test		0.078	19:15
Air Blank		0.000	19:15
Subject Sample	# 1	0.064	19:17
Breath Volume		2.246	19:17
Air Blank		0.000	19:18
Air Blank		0.000	19:20
Subject Sample	# 2	0.056	19:20
Breath Volume		1.890	19:20
Air Blank		0.000	19:21
Control Test		0.077	19:21
Air Blank		0.000	19:22
Diagnostics Check		OK	19:22

Data Download Date/Time: 12/20/2017 11:34

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITE,RYAN,P
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: BISSOON,STEPHEN
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A1
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:58

SUBJECT'S LAST NAME: RICHARDS

DATE/TIME OF TEST: 11/28/2017 20:32:58

Diagnostics Check		OK	20:24
Air Blank		0.000	20:24
Control Test		0.077	20:24
Air Blank		0.000	20:25
Subject Sample	# 1	0.228	20:27
Breath Volume		1.363	20:27
Air Blank		0.000	20:28
Air Blank		0.000	20:30
Subject Sample	# 2	0.227	20:31
Breath Volume		1.800	20:31
Air Blank		0.000	20:32
Control Test		0.076	20:32
Air Blank		0.000	20:32
Diagnostics Check		OK	20:32

Data Download Date/Time: 12/20/2017 11:34

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,BRYAN,C
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: RENTERIA,BARTOLO
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A1
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: GUERRERO-PENA

DATE/TIME OF TEST: 11/29/2017 02:38:19

Diagnostics Check		OK	02:35	
Air Blank		0.000	02:35	
Control Test		0.077	02:36	
Air Blank		0.000	02:36	
Subject Sample	# 1		02:36	Subject Test Refused
Breath Volume		0.000	02:36	
Air Blank		0.000	02:37	
Control Test		0.078	02:37	
Air Blank		0.000	02:38	
Diagnostics Check		OK	02:38	

Data Download Date/Time: 12/20/2017 11:34

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006633

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANGELICCHI,JOSEPH,S
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: ANGELICCHI,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A3
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:20

SUBJECT'S LAST NAME: WRIGHT

DATE/TIME OF TEST: 11/14/2017 20:46:46

Diagnostics Check		OK	20:40
Air Blank		0.000	20:40
Control Test		0.079	20:40
Air Blank		0.000	20:41
Subject Sample	# 1	0.116	20:41
Breath Volume		4.136	20:41
Air Blank		0.000	20:42
Air Blank		0.000	20:44
Subject Sample	# 2	0.120	20:44
Breath Volume		3.625	20:44
Air Blank		0.000	20:45
Control Test		0.079	20:46
Air Blank		0.000	20:46
Diagnostics Check		OK	20:46

Data Download Date/Time: 12/28/2017 12:21

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GWODZ,OFC,R
BREATH TEST OPERATOR AGENCY: Bradenton PD
ARRESTING OFFICER: GWODZ,OFC
ARRESTING OFFICER AGENCY: Bradenton PD
LAST AGENCY INSPECTION: 11/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 21:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:52

SUBJECT'S LAST NAME: DISBERRY

DATE/TIME OF TEST: 11/29/2017 22:23:14

Diagnostics Check		OK	22:15
Air Blank		0.000	22:15
Control Test		0.079	22:16
Air Blank		0.000	22:16
Subject Sample	# 1	0.216	22:18
Breath Volume		1.945	22:18
Air Blank		0.000	22:19
Air Blank		0.000	22:20
Subject Sample	# 2	0.220	22:21
Breath Volume		2.355	22:21
Air Blank		0.000	22:22
Control Test		0.077	22:22
Air Blank		0.000	22:23
Diagnostics Check		OK	22:23

Data Download Date/Time: 12/18/2017 11:05

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006636

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HITCHCOCK,WILLIAM,D
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: HITCHCOCK,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 10/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A3
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:52

SUBJECT'S LAST NAME: TELEBRICO

DATE/TIME OF TEST: 11/26/2017 02:19:04

Diagnostics Check		OK	02:13	
Air Blank		0.000	02:14	
Control Test		0.079	02:14	
Air Blank		0.000	02:14	
Subject Sample	# 1	0.210	02:15	
Breath Volume		2.472	02:15	
Air Blank		0.000	02:15	
Air Blank		0.000	02:17	
Subject Sample	# 2	0.209	02:18	
Breath Volume		2.929	02:18	
Air Blank			02:18	RFI Detect
Air Blank		0.000	02:19	

Data Download Date/Time: 11/28/2017 04:27

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006636

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HITCHCOCK,WILLIAM,D
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: HITCHCOCK,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 10/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A3
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:52

SUBJECT'S LAST NAME: TELEBRICO

DATE/TIME OF TEST: 11/26/2017 02:27:31

Diagnostics Check		OK	02:21
Air Blank		0.000	02:21
Control Test		0.079	02:21
Air Blank		0.000	02:22
Subject Sample	# 1	0.214	02:22
Breath Volume		2.710	02:22
Air Blank		0.000	02:23
Air Blank		0.000	02:25
Subject Sample	# 2	0.216	02:25
Breath Volume		2.882	02:25
Air Blank		0.000	02:26
Control Test		0.079	02:26
Air Blank		0.000	02:27
Diagnostics Check		OK	02:27

Data Download Date/Time: 11/28/2017 04:27

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006637

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KASELAK,STEPHEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/03/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A3
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:07

SUBJECT'S LAST NAME: KEY

DATE/TIME OF TEST: 11/04/2017 01:38:59

Diagnostics Check	OK	01:36	
Air Blank	0.000	01:37	
Control Test	0.077	01:37	
Air Blank		01:38	Purge Fail
Air Blank	0.000	01:38	

Data Download Date/Time: 11/12/2017 11:52

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006637

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KASELAK,STEPHEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/03/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A3
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:07

SUBJECT'S LAST NAME: KEY

DATE/TIME OF TEST: 11/04/2017 01:42:14

Diagnostics Check	OK	01:40	
Air Blank	0.000	01:41	
Control Test	0.077	01:41	Interferent Detect
Air Blank	0.000	01:42	

Data Download Date/Time: 11/12/2017 11:52

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006637

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KASELAK,STEPHEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/03/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A3
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:07

SUBJECT'S LAST NAME: KEY

DATE/TIME OF TEST: 11/04/2017 01:52:14

Diagnostics Check		OK	01:45
Air Blank		0.000	01:45
Control Test		0.079	01:46
Air Blank		0.000	01:47
Subject Sample	# 1	0.122	01:47
Breath Volume		3.347	01:47
Air Blank		0.000	01:48
Air Blank		0.000	01:50
Subject Sample	# 2	0.121	01:50
Breath Volume		3.074	01:50
Air Blank		0.000	01:51
Control Test		0.081	01:51
Air Blank		0.000	01:52
Diagnostics Check		OK	01:52

Data Download Date/Time: 11/12/2017 11:52

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006637

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANGELICCHI,JOSEPH,S
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: ANGELICCHI,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 10/03/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A3
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:09

SUBJECT'S LAST NAME: MCFARLANE

DATE/TIME OF TEST: 11/04/2017 02:38:25

Diagnostics Check		OK	02:31
Air Blank		0.000	02:31
Control Test		0.081	02:32
Air Blank		0.000	02:32
Subject Sample	# 1	0.177	02:33
Breath Volume		1.550	02:33
Air Blank		0.000	02:34
Air Blank		0.000	02:35
Subject Sample	# 2	0.171	02:36
Breath Volume		1.601	02:36
Air Blank		0.000	02:37
Control Test		0.080	02:37
Air Blank		0.000	02:38
Diagnostics Check		OK	02:38

Data Download Date/Time: 11/12/2017 11:52

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006692

INSTRUMENT REGISTERED TO: HOLMES BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORES,ALAN,P
BREATH TEST OPERATOR AGENCY: Holmes Beach PD
ARRESTING OFFICER: BORES,ALAN
ARRESTING OFFICER AGENCY: Holmes Beach PD
LAST AGENCY INSPECTION: 10/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A2
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:08

SUBJECT'S LAST NAME: WEBER

DATE/TIME OF TEST: 11/22/2017 04:36:45

Diagnostics Check		OK	04:29
Air Blank		0.000	04:30
Control Test		0.078	04:30
Air Blank		0.000	04:30
Subject Sample	# 1	0.140	04:31
Breath Volume		1.687	04:31
Air Blank		0.000	04:32
Air Blank		0.000	04:34
Subject Sample	# 2	0.143	04:35
Breath Volume		2.214	04:35
Air Blank		0.000	04:35
Control Test		0.078	04:36
Air Blank		0.000	04:36
Diagnostics Check		OK	04:36

Data Download Date/Time: 11/27/2017 10:17

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006692

INSTRUMENT REGISTERED TO: HOLMES BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORES,ALAN,P
BREATH TEST OPERATOR AGENCY: Holmes Beach PD
ARRESTING OFFICER: BORES,ALAN
ARRESTING OFFICER AGENCY: Holmes Beach PD
LAST AGENCY INSPECTION: 10/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A2
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:27

SUBJECT'S LAST NAME: RILEY

DATE/TIME OF TEST: 11/24/2017 00:56:06

Diagnostics Check		OK	00:49
Air Blank		0.000	00:50
Control Test		0.078	00:50
Air Blank		0.000	00:51
Subject Sample	# 1	0.000	00:51
Breath Volume		1.535	00:51
Air Blank		0.000	00:52
Air Blank		0.000	00:54
Subject Sample	# 2	0.000	00:54
Breath Volume		1.667	00:54
Air Blank		0.000	00:55
Control Test		0.078	00:55
Air Blank		0.000	00:56
Diagnostics Check		OK	00:56

Data Download Date/Time: 11/27/2017 10:17

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PESTKA,PAIGE,R
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: HOPPE,ZACH
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 11/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 21:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:18

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 11/16/2017 22:52:49

Diagnostics Check		OK	22:41	
Air Blank		0.000	22:41	
Control Test		0.077	22:42	
Air Blank		0.000	22:42	
Subject Sample	# 1		22:45	No Sample Provided
Breath Volume		0.000	22:45	
Air Blank		0.000	22:46	
Air Blank		0.000	22:48	
Subject Sample	# 2		22:51	No Sample Provided
Breath Volume		0.000	22:51	
Air Blank		0.000	22:51	
Control Test		0.079	22:52	
Air Blank		0.000	22:52	
Diagnostics Check		OK	22:52	

Data Download Date/Time: 12/21/2017 15:07

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,SHAWN,W
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: MARTIN,SHAWN
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 11/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 20:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:05

SUBJECT'S LAST NAME: LADDEN

DATE/TIME OF TEST: 11/17/2017 21:32:08

Diagnostics Check		OK	21:25
Air Blank		0.000	21:26
Control Test		0.079	21:26
Air Blank		0.000	21:26
Subject Sample	# 1	0.137	21:27
Breath Volume		1.285	21:27
Air Blank		0.000	21:28
Air Blank		0.000	21:30
Subject Sample	# 2	0.148	21:30
Breath Volume		2.468	21:30
Air Blank		0.000	21:31
Control Test		0.078	21:31
Air Blank		0.000	21:32
Diagnostics Check		OK	21:32

Data Download Date/Time: 12/21/2017 15:07

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HODGES,TITUS,R
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: LARIOS,JORGE
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 11/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 22:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: WATNEE

DATE/TIME OF TEST: 11/17/2017 23:19:34

Diagnostics Check		OK	23:10
Air Blank		0.000	23:11
Control Test		0.079	23:11
Air Blank		0.000	23:12
Subject Sample	# 1	0.162	23:14
Breath Volume		1.136	23:14
Air Blank		0.000	23:14
Air Blank		0.000	23:16
Subject Sample	# 2	0.174	23:17
Breath Volume		2.894	23:17
Air Blank		0.000	23:18
Control Test		0.078	23:18
Air Blank		0.000	23:19
Diagnostics Check		OK	23:19

Data Download Date/Time: 12/21/2017 15:07

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALVAREZ,ORLANDO,
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: ALVAREZ,ORLANDO
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 11/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 14:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:23

SUBJECT'S LAST NAME: AVIS

DATE/TIME OF TEST: 11/22/2017 14:47:26

Diagnostics Check		OK	14:44	
Air Blank		0.000	14:44	
Control Test		0.078	14:45	
Air Blank		0.000	14:45	
Subject Sample	# 1		14:46	Subject Test Refused
Breath Volume		0.000	14:46	
Air Blank		0.000	14:46	
Control Test		0.078	14:46	
Air Blank		0.000	14:47	
Diagnostics Check		OK	14:47	

Data Download Date/Time: 12/21/2017 15:07

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAEZ,HAROLD,
BREATH TEST OPERATOR AGENCY: Palmetto PD
ARRESTING OFFICER: BAEZ,HAROLD
ARRESTING OFFICER AGENCY: Palmetto PD
LAST AGENCY INSPECTION: 10/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 05:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:05

SUBJECT'S LAST NAME: NODA

DATE/TIME OF TEST: 11/01/2017 06:35:05

Diagnostics Check		OK	06:28
Air Blank		0.000	06:29
Control Test		0.080	06:29
Air Blank		0.000	06:30
Subject Sample	# 1	0.122	06:30
Breath Volume		1.187	06:30
Air Blank		0.000	06:31
Air Blank		0.000	06:33
Subject Sample	# 2	0.126	06:33
Breath Volume		2.082	06:33
Air Blank		0.000	06:34
Control Test		0.080	06:34
Air Blank		0.000	06:35
Diagnostics Check		OK	06:35

Data Download Date/Time: 11/27/2017 09:58

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 10/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 17:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:02

SUBJECT'S LAST NAME: RANAGAN

DATE/TIME OF TEST: 11/03/2017 18:29:11

Diagnostics Check		OK	18:23
Air Blank		0.000	18:23
Control Test		0.080	18:23
Air Blank		0.000	18:24
Subject Sample	# 1	0.114	18:24
Breath Volume		1.582	18:24
Air Blank		0.000	18:25
Air Blank		0.000	18:27
Subject Sample	# 2	0.115	18:27
Breath Volume		2.550	18:27
Air Blank		0.000	18:28
Control Test		0.078	18:28
Air Blank		0.000	18:29
Diagnostics Check		OK	18:29

Data Download Date/Time: 11/27/2017 09:58

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUICK,BENJAMIN,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: QUICK,BENJAMIN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 10/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 22:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 11/04/2017 23:18:36

Diagnostics Check		OK	23:15	
Air Blank		0.000	23:16	
Control Test		0.081	23:16	
Air Blank		0.000	23:17	
Subject Sample	# 1		23:17	Subject Test Refused
Breath Volume		0.000	23:17	
Air Blank		0.000	23:17	
Control Test		0.080	23:17	
Air Blank		0.000	23:18	
Diagnostics Check		OK	23:18	

Data Download Date/Time: 11/27/2017 09:58

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 10/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 23:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:34

SUBJECT'S LAST NAME: BASILIO

DATE/TIME OF TEST: 11/05/2017 00:12:08

Diagnostics Check		OK	23:58	
Air Blank		0.000	23:59	
Control Test		0.079	23:59	
Air Blank		0.000	00:00	
Subject Sample	# 1	0.175	00:01	
Breath Volume		1.597	00:01	
Air Blank		0.000	00:02	
Air Blank		0.000	00:03	
Subject Sample	# 2	0.155	00:07	Volume Not Met
Breath Volume		0.058	00:07	
Air Blank		0.000	00:08	
Air Blank		0.000	00:09	
Subject Sample	# 3		00:10	Subject Test Refused
Breath Volume		0.039	00:10	
Air Blank		0.000	00:11	
Control Test		0.073	00:11	Control Outside Tolerance
Air Blank		0.000	00:12	

Data Download Date/Time: 11/27/2017 09:58

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUICK,BENJAMIN,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: QUICK,BENJAMIN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 10/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 00:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:34

SUBJECT'S LAST NAME: WILLERS

DATE/TIME OF TEST: 11/05/2017 01:16:32

Diagnostics Check		OK	01:10
Air Blank		0.000	01:10
Control Test		0.080	01:10
Air Blank		0.000	01:11
Subject Sample	# 1	0.064	01:12
Breath Volume		1.769	01:12
Air Blank		0.000	01:12
Air Blank		0.000	01:14
Subject Sample	# 2	0.061	01:15
Breath Volume		3.796	01:15
Air Blank		0.000	01:15
Control Test		0.080	01:15
Air Blank		0.000	01:16
Diagnostics Check		OK	01:16

Data Download Date/Time: 11/27/2017 09:58

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 10/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: VINCENT

DATE/TIME OF TEST: 11/05/2017 01:25:32

Diagnostics Check		OK	01:22	
Air Blank		0.000	01:23	
Control Test		0.079	01:23	
Air Blank		0.000	01:24	
Subject Sample	# 1		01:24	Subject Test Refused
Breath Volume		0.000	01:24	
Air Blank		0.000	01:24	
Control Test		0.079	01:24	
Air Blank		0.000	01:25	
Diagnostics Check		OK	01:25	

Data Download Date/Time: 11/27/2017 09:58

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUICK,BENJAMIN,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: QUICK,BENJAMIN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 10/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 18:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:35

SUBJECT'S LAST NAME: FOULDS

DATE/TIME OF TEST: 11/05/2017 20:03:00

Diagnostics Check		OK	19:59	
Air Blank		0.000	19:59	
Control Test		0.079	19:59	
Air Blank		0.000	20:00	
Subject Sample	# 1		20:01	Subject Test Refused
Breath Volume		0.000	20:01	
Air Blank		0.000	20:02	
Control Test		0.080	20:02	
Air Blank		0.000	20:02	
Diagnostics Check		OK	20:03	

Data Download Date/Time: 11/27/2017 09:58

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WATSON,DANIEL,L
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: WATSON,DANIEL
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 10/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 14:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:00

SUBJECT'S LAST NAME: PANIAGUA

DATE/TIME OF TEST: 11/06/2017 18:44:27

Diagnostics Check		OK	18:38
Air Blank		0.000	18:38
Control Test		0.078	18:39
Air Blank		0.000	18:39
Subject Sample	# 1	0.267	18:39
Breath Volume		1.296	18:39
Air Blank		0.000	18:40
Air Blank		0.000	18:42
Subject Sample	# 2	0.262	18:42
Breath Volume		1.441	18:42
Air Blank		0.000	18:43
Control Test		0.076	18:43
Air Blank		0.000	18:44
Diagnostics Check		OK	18:44

Data Download Date/Time: 11/27/2017 09:58

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KENYAN,M,A
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: KENYAN,MA
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 10/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 19:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:13

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 11/06/2017 20:41:16

Diagnostics Check		OK	20:38	
Air Blank		0.000	20:38	
Control Test		0.080	20:39	
Air Blank		0.000	20:39	
Subject Sample	# 1		20:39	Subject Test Refused
Breath Volume		0.000	20:39	
Air Blank		0.000	20:40	
Control Test		0.079	20:40	
Air Blank		0.000	20:41	
Diagnostics Check		OK	20:41	

Data Download Date/Time: 11/27/2017 09:58

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORES,ALAN,P
BREATH TEST OPERATOR AGENCY: Holmes Beach PD
ARRESTING OFFICER: BORES,ALAN
ARRESTING OFFICER AGENCY: Holmes Beach PD
LAST AGENCY INSPECTION: 10/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 01:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: DORRIAN

DATE/TIME OF TEST: 11/07/2017 03:01:03

Diagnostics Check		OK	02:54
Air Blank		0.000	02:54
Control Test		0.080	02:55
Air Blank		0.000	02:55
Subject Sample	# 1	0.172	02:56
Breath Volume		2.855	02:56
Air Blank		0.000	02:57
Air Blank		0.000	02:58
Subject Sample	# 2	0.164	02:59
Breath Volume		3.296	02:59
Air Blank		0.000	03:00
Control Test		0.079	03:00
Air Blank		0.000	03:00
Diagnostics Check		OK	03:01

Data Download Date/Time: 11/27/2017 09:58

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KENYAN,M,A
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: KENYAN,MA
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 10/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 02:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:02

SUBJECT'S LAST NAME: RUDOLPH

DATE/TIME OF TEST: 11/08/2017 03:29:46

Diagnostics Check		OK	03:23
Air Blank		0.000	03:23
Control Test		0.080	03:23
Air Blank		0.000	03:24
Subject Sample	# 1	0.166	03:25
Breath Volume		1.765	03:25
Air Blank		0.000	03:25
Air Blank		0.000	03:27
Subject Sample	# 2	0.180	03:28
Breath Volume		1.160	03:28
Air Blank		0.000	03:28
Control Test		0.079	03:29
Air Blank		0.000	03:29
Diagnostics Check		OK	03:29

Data Download Date/Time: 11/27/2017 09:59

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORES,ALAN,P
BREATH TEST OPERATOR AGENCY: Holmes Beach PD
ARRESTING OFFICER: BORES,ALAN
ARRESTING OFFICER AGENCY: Holmes Beach PD
LAST AGENCY INSPECTION: 10/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 03:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:34

SUBJECT'S LAST NAME: EGAN

DATE/TIME OF TEST: 11/08/2017 05:00:19

Diagnostics Check		OK	04:56	
Air Blank		0.000	04:56	
Control Test		0.080	04:57	
Air Blank		0.000	04:57	
Subject Sample	# 1		04:58	Subject Test Refused
Breath Volume		0.000	04:58	
Air Blank		0.000	04:59	
Control Test		0.080	04:59	
Air Blank		0.000	05:00	
Diagnostics Check		OK	05:00	

Data Download Date/Time: 11/27/2017 09:59

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORES,ALAN,P
BREATH TEST OPERATOR AGENCY: Holmes Beach PD
ARRESTING OFFICER: BORES,ALAN
ARRESTING OFFICER AGENCY: Holmes Beach PD
LAST AGENCY INSPECTION: 10/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 03:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:34

SUBJECT'S LAST NAME: EGAN

DATE/TIME OF TEST: 11/08/2017 05:09:47

Diagnostics Check		OK	05:02
Air Blank		0.000	05:03
Control Test		0.080	05:03
Air Blank		0.000	05:04
Subject Sample	# 1	0.214	05:04
Breath Volume		2.175	05:04
Air Blank		0.000	05:05
Air Blank		0.000	05:07
Subject Sample	# 2	0.217	05:08
Breath Volume		1.546	05:08
Air Blank		0.000	05:08
Control Test		0.079	05:09
Air Blank		0.000	05:09
Diagnostics Check		OK	05:09

Data Download Date/Time: 11/27/2017 09:59

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 10/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 18:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:12

SUBJECT'S LAST NAME: PETERSON

DATE/TIME OF TEST: 11/09/2017 19:38:26

Diagnostics Check		OK	19:32
Air Blank		0.000	19:32
Control Test		0.081	19:33
Air Blank		0.000	19:33
Subject Sample	# 1	0.042	19:34
Breath Volume		1.292	19:34
Air Blank		0.000	19:34
Air Blank		0.000	19:36
Subject Sample	# 2	0.034	19:36
Breath Volume		1.121	19:36
Air Blank		0.000	19:37
Control Test		0.078	19:37
Air Blank		0.000	19:38
Diagnostics Check		OK	19:38

Data Download Date/Time: 11/27/2017 09:59

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUICK,BENJAMIN,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: QUICK,BENJAMIN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 10/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 20:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:17

SUBJECT'S LAST NAME: HENRY RAMOS

DATE/TIME OF TEST: 11/13/2017 21:48:14

Diagnostics Check		OK	21:35	
Air Blank		0.000	21:36	
Control Test		0.080	21:36	
Air Blank		0.000	21:36	
Subject Sample	# 1	0.175	21:40	Volume Not Met
Breath Volume		0.289	21:40	
Air Blank		0.000	21:40	
Air Blank		0.000	21:42	
Subject Sample	# 2	0.226	21:43	
Breath Volume		2.273	21:43	
Air Blank		0.000	21:44	
Air Blank		0.000	21:46	
Subject Sample	# 3	0.232	21:46	
Breath Volume		3.046	21:46	
Air Blank		0.000	21:47	
Control Test		0.079	21:47	
Air Blank		0.000	21:48	
Diagnostics Check		OK	21:48	

Data Download Date/Time: 11/27/2017 09:59

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DONALD,JOHNATHAN,K
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: ANGELICCHI,JOE
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 10/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 02:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: THORESON

DATE/TIME OF TEST: 11/14/2017 04:33:00

Diagnostics Check		OK	04:26
Air Blank		0.000	04:26
Control Test		0.080	04:27
Air Blank		0.000	04:27
Subject Sample	# 1	0.202	04:28
Breath Volume		3.703	04:28
Air Blank		0.000	04:29
Air Blank		0.000	04:30
Subject Sample	# 2	0.192	04:31
Breath Volume		2.246	04:31
Air Blank		0.000	04:32
Control Test		0.076	04:32
Air Blank		0.000	04:32
Diagnostics Check		OK	04:33

Data Download Date/Time: 11/27/2017 09:59

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAKAI,MICHAEL,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: MAKAI,MICHAEL
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 10/24/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 21:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:46

SUBJECT'S LAST NAME: RUIZ

DATE/TIME OF TEST: 11/14/2017 22:12:45

Diagnostics Check		OK	22:06
Air Blank		0.000	22:07
Control Test		0.080	22:07
Air Blank		0.000	22:08
Subject Sample	# 1	0.000	22:08
Breath Volume		1.847	22:08
Air Blank		0.000	22:08
Air Blank		0.000	22:10
Subject Sample	# 2	0.000	22:11
Breath Volume		1.511	22:11
Air Blank		0.000	22:11
Control Test		0.079	22:12
Air Blank		0.000	22:12
Diagnostics Check		OK	22:12

Data Download Date/Time: 11/27/2017 09:59

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 10/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 18:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:37

SUBJECT'S LAST NAME: BRATE III

DATE/TIME OF TEST: 11/17/2017 19:10:12

Diagnostics Check		OK	19:07	
Air Blank		0.000	19:07	
Control Test		0.080	19:08	
Air Blank		0.000	19:08	
Subject Sample	# 1		19:08	Subject Test Refused
Breath Volume		0.000	19:08	
Air Blank		0.000	19:09	
Control Test		0.081	19:09	
Air Blank		0.000	19:10	
Diagnostics Check		OK	19:10	

Data Download Date/Time: 11/27/2017 09:59

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUICK,BENJAMIN,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: QUICK,BENJAMIN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 10/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 20:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:51

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 11/17/2017 21:17:32

Diagnostics Check		OK	21:11
Air Blank		0.000	21:11
Control Test		0.080	21:12
Air Blank		0.000	21:12
Subject Sample	# 1	0.220	21:13
Breath Volume		2.605	21:13
Air Blank		0.000	21:13
Air Blank		0.000	21:15
Subject Sample	# 2	0.206	21:15
Breath Volume		2.500	21:15
Air Blank		0.000	21:16
Control Test		0.080	21:16
Air Blank		0.000	21:17
Diagnostics Check		OK	21:17

Data Download Date/Time: 11/27/2017 09:59

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DONALD,JOHNATHAN,K
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: DONALD,JOHNATHAN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 10/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 20:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:10

SUBJECT'S LAST NAME: BARTOLON PEREZ

DATE/TIME OF TEST: 11/18/2017 21:54:57

Diagnostics Check		OK	21:46
Air Blank		0.000	21:47
Control Test		0.080	21:47
Air Blank		0.000	21:48
Subject Sample	# 1	0.200	21:49
Breath Volume		1.210	21:49
Air Blank		0.000	21:50
Air Blank		0.000	21:52
Subject Sample	# 2	0.197	21:53
Breath Volume		1.855	21:53
Air Blank		0.000	21:54
Control Test		0.076	21:54
Air Blank		0.000	21:54
Diagnostics Check		OK	21:54

Data Download Date/Time: 11/27/2017 09:59

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUICK,BENJAMIN,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: QUICK,BENJAMIN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 10/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 22:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: GIELAZYN

DATE/TIME OF TEST: 11/18/2017 23:13:00

Diagnostics Check		OK	23:10	
Air Blank		0.000	23:10	
Control Test		0.081	23:10	
Air Blank		0.000	23:11	
Subject Sample	# 1		23:11	Subject Test Refused
Breath Volume		0.000	23:11	
Air Blank		0.000	23:12	
Control Test		0.080	23:12	
Air Blank		0.000	23:12	
Diagnostics Check		OK	23:13	

Data Download Date/Time: 11/27/2017 09:59

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 10/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 00:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 11/19/2017 01:24:21

Diagnostics Check		OK	01:21	
Air Blank		0.000	01:21	
Control Test		0.081	01:22	
Air Blank		0.000	01:22	
Subject Sample	# 1		01:22	Subject Test Refused
Breath Volume		0.000	01:22	
Air Blank		0.000	01:23	
Control Test		0.080	01:23	
Air Blank		0.000	01:24	
Diagnostics Check		OK	01:24	

Data Download Date/Time: 11/27/2017 09:59

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAKAI,MICHAEL,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: MAKAI,MICHAEL
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 10/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 00:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 11/19/2017 02:20:00

Diagnostics Check		OK	02:10
Air Blank		0.000	02:11
Control Test		0.081	02:11
Air Blank		0.000	02:12
Subject Sample	# 1	0.205	02:15
Breath Volume		2.195	02:15
Air Blank		0.000	02:16
Air Blank		0.000	02:17
Subject Sample	# 2	0.206	02:18
Breath Volume		2.492	02:18
Air Blank		0.000	02:19
Control Test		0.079	02:19
Air Blank		0.000	02:19
Diagnostics Check		OK	02:20

Data Download Date/Time: 11/27/2017 09:59

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUICK,BENJAMIN,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: QUICK,BENJAMIN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 10/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 02:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:49

SUBJECT'S LAST NAME: BLAIR

DATE/TIME OF TEST: 11/19/2017 03:25:20

Diagnostics Check		OK	03:21	
Air Blank		0.000	03:22	
Control Test		0.081	03:22	
Air Blank		0.000	03:23	
Subject Sample	# 1		03:23	Subject Test Refused
Breath Volume		0.000	03:23	
Air Blank		0.000	03:24	
Control Test		0.081	03:24	
Air Blank		0.000	03:25	
Diagnostics Check		OK	03:25	

Data Download Date/Time: 11/27/2017 09:59

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUICK,BENJAMIN,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: QUICK,BENJAMIN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 10/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 01:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:32

SUBJECT'S LAST NAME: MENDOZA

DATE/TIME OF TEST: 11/20/2017 02:59:24

Diagnostics Check		OK	02:53
Air Blank		0.000	02:53
Control Test		0.081	02:53
Air Blank		0.000	02:54
Subject Sample	# 1	0.248	02:54
Breath Volume		2.636	02:54
Air Blank		0.000	02:55
Air Blank		0.000	02:57
Subject Sample	# 2	0.243	02:57
Breath Volume		3.042	02:57
Air Blank		0.000	02:58
Control Test		0.080	02:58
Air Blank		0.000	02:59
Diagnostics Check		OK	02:59

Data Download Date/Time: 11/27/2017 09:59

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMPSON,STEVEN,M
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: SIMPSON,STEVEN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 10/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 02:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:54

SUBJECT'S LAST NAME: JORDAN

DATE/TIME OF TEST: 11/21/2017 04:20:46

Diagnostics Check		OK	04:14
Air Blank		0.000	04:15
Control Test		0.081	04:15
Air Blank		0.000	04:15
Subject Sample	# 1	0.151	04:16
Breath Volume		3.113	04:16
Air Blank		0.000	04:17
Air Blank		0.000	04:18
Subject Sample	# 2	0.146	04:19
Breath Volume		2.273	04:19
Air Blank		0.000	04:19
Control Test		0.080	04:20
Air Blank		0.000	04:20
Diagnostics Check		OK	04:20

Data Download Date/Time: 11/27/2017 09:59

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUICK,BENJAMIN,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: QUICK,BENJAMIN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 10/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 00:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:07

SUBJECT'S LAST NAME: CRUZ MENDOZA

DATE/TIME OF TEST: 11/23/2017 01:42:54

Diagnostics Check		OK	01:27	
Air Blank		0.000	01:28	
Control Test		0.081	01:28	
Air Blank		0.000	01:29	
Subject Sample	# 1	0.091	01:32	Volume Not Met
Breath Volume		1.003	01:32	
Air Blank		0.000	01:33	
Air Blank		0.000	01:34	
Subject Sample	# 2	0.121	01:36	
Breath Volume		2.078	01:36	
Air Blank		0.000	01:36	
Air Blank		0.000	01:38	
Subject Sample	# 3	0.123	01:41	
Breath Volume		2.730	01:41	
Air Blank		0.000	01:41	
Control Test		0.080	01:42	
Air Blank		0.000	01:42	
Diagnostics Check		OK	01:42	

Data Download Date/Time: 11/27/2017 09:59

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAKAI,MICHAEL,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: MAKAI,MICHAEL
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 10/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:36

SUBJECT'S LAST NAME: HAWKINS

DATE/TIME OF TEST: 11/23/2017 04:05:57

Diagnostics Check		OK	03:57
Air Blank		0.000	03:57
Control Test		0.080	03:57
Air Blank		0.000	03:58
Subject Sample	# 1	0.185	04:00
Breath Volume		1.257	04:00
Air Blank		0.000	04:01
Air Blank		0.000	04:03
Subject Sample	# 2	0.173	04:04
Breath Volume		1.167	04:04
Air Blank		0.000	04:05
Control Test		0.079	04:05
Air Blank		0.000	04:05
Diagnostics Check		OK	04:05

Data Download Date/Time: 11/27/2017 09:59

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORES,ALAN,P
BREATH TEST OPERATOR AGENCY: Holmes Beach PD
ARRESTING OFFICER: BORES,ALAN
ARRESTING OFFICER AGENCY: Holmes Beach PD
LAST AGENCY INSPECTION: 10/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: HAGELE

DATE/TIME OF TEST: 11/26/2017 02:59:35

Diagnostics Check		OK	02:56	
Air Blank		0.000	02:57	
Control Test		0.080	02:57	
Air Blank		0.000	02:58	
Subject Sample	# 1		02:58	Subject Test Refused
Breath Volume		0.000	02:58	
Air Blank		0.000	02:58	
Control Test		0.081	02:58	
Air Blank		0.000	02:59	
Diagnostics Check		OK	02:59	

Data Download Date/Time: 11/27/2017 09:59

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMPSON,STEVEN,M
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: SIMPSON,STEVEN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 10/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 19:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: PULLAM

DATE/TIME OF TEST: 11/26/2017 20:33:53

Diagnostics Check		OK	20:23
Air Blank		0.000	20:24
Control Test		0.081	20:24
Air Blank		0.000	20:25
Subject Sample	# 1	0.222	20:25
Breath Volume		2.367	20:25
Air Blank		0.000	20:26
Air Blank		0.000	20:28
Subject Sample	# 2	0.248	20:29
Breath Volume		2.144	20:29
Air Blank		0.000	20:29
Air Blank		0.000	20:31
Subject Sample	# 3	0.235	20:32
Breath Volume		2.324	20:32
Air Blank		0.000	20:32
Control Test		0.079	20:33
Air Blank		0.000	20:33
Diagnostics Check		OK	20:33

Data Download Date/Time: 11/27/2017 09:59

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUICK,BENJAMIN,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: QUICK,BENJAMIN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 11/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 19:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:30

SUBJECT'S LAST NAME: TABBITA

DATE/TIME OF TEST: 11/27/2017 21:03:01

Diagnostics Check		OK	21:00	
Air Blank		0.000	21:00	
Control Test		0.080	21:00	
Air Blank		0.000	21:01	
Subject Sample	# 1		21:01	Subject Test Refused
Breath Volume		0.000	21:01	
Air Blank		0.000	21:02	
Control Test		0.080	21:02	
Air Blank		0.000	21:02	
Diagnostics Check		OK	21:03	

Data Download Date/Time: 12/18/2017 12:40

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUICK,BENJAMIN,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: QUICK,BENJAMIN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 11/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 00:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:33

SUBJECT'S LAST NAME: RAPSON

DATE/TIME OF TEST: 11/29/2017 00:56:57

Diagnostics Check		OK	00:54	
Air Blank		0.000	00:54	
Control Test		0.081	00:54	
Air Blank		0.000	00:55	
Subject Sample	# 1		00:55	RFI Detect
Breath Volume		0.000	00:55	
Air Blank		0.000	00:56	
Control Test		0.081	00:56	
Air Blank		0.000	00:56	
Diagnostics Check		OK	00:56	

Data Download Date/Time: 12/18/2017 12:40

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUICK,BENJAMIN,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: QUICK,BENJAMIN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 11/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 00:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:33

SUBJECT'S LAST NAME: RAPSON

DATE/TIME OF TEST: 11/29/2017 01:04:36

Diagnostics Check		OK	00:58
Air Blank		0.000	00:58
Control Test		0.080	00:59
Air Blank		0.000	00:59
Subject Sample	# 1	0.104	01:00
Breath Volume		1.386	01:00
Air Blank		0.000	01:00
Air Blank		0.000	01:02
Subject Sample	# 2	0.101	01:03
Breath Volume		2.472	01:03
Air Blank		0.000	01:03
Control Test		0.081	01:03
Air Blank		0.000	01:04
Diagnostics Check		OK	01:04

Data Download Date/Time: 12/18/2017 12:40

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006758

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 10/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 123456789
DRY GAS EXPIRY DATE: 11/11/2018
VIOLATION CODE: DUI
ARREST TIME: 02:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: STEPHENS

DATE/TIME OF TEST: 11/11/2017 03:13:51

Diagnostics Check		OK	03:06
Air Blank		0.000	03:07
Control Test		0.080	03:07
Air Blank		0.000	03:08
Subject Sample	# 1	0.222	03:09
Breath Volume		2.609	03:09
Air Blank		0.000	03:10
Air Blank		0.000	03:11
Subject Sample	# 2	0.238	03:12
Breath Volume		2.648	03:12
Air Blank		0.000	03:12
Control Test		0.079	03:13
Air Blank		0.000	03:13
Diagnostics Check		OK	03:13

Data Download Date/Time: 11/27/2017 10:15

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006758

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 10/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 123456789
DRY GAS EXPIRY DATE: 11/11/2018
VIOLATION CODE: DUI
ARREST TIME: 20:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:35

SUBJECT'S LAST NAME: LOPEZ HERNANDEZ

DATE/TIME OF TEST: 11/11/2017 21:04:40

Diagnostics Check		OK	20:57
Air Blank		0.000	20:58
Control Test		0.080	20:58
Air Blank		0.000	20:59
Subject Sample	# 1	0.138	20:59
Breath Volume		1.906	20:59
Air Blank		0.000	21:00
Air Blank		0.000	21:02
Subject Sample	# 2	0.145	21:02
Breath Volume		2.808	21:02
Air Blank		0.000	21:03
Control Test		0.080	21:04
Air Blank		0.000	21:04
Diagnostics Check		OK	21:04

Data Download Date/Time: 11/27/2017 10:15

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006758

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 10/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 123456789
DRY GAS EXPIRY DATE: 11/11/2018
VIOLATION CODE: DUI
ARREST TIME: 00:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: PINEDA RAMOS

DATE/TIME OF TEST: 11/13/2017 01:32:54

Diagnostics Check		OK	01:25
Air Blank		0.000	01:25
Control Test		0.080	01:26
Air Blank		0.000	01:26
Subject Sample	# 1	0.212	01:27
Breath Volume		1.777	01:27
Air Blank		0.000	01:28
Air Blank		0.000	01:30
Subject Sample	# 2	0.207	01:31
Breath Volume		1.687	01:31
Air Blank		0.000	01:31
Control Test		0.080	01:32
Air Blank		0.000	01:32
Diagnostics Check		OK	01:32

Data Download Date/Time: 11/27/2017 10:15

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006758

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 10/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 123456789
DRY GAS EXPIRY DATE: 11/11/2018
VIOLATION CODE: DUI
ARREST TIME: 00:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:38

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 11/25/2017 01:07:50

Diagnostics Check		OK	01:00
Air Blank		0.000	01:01
Control Test		0.080	01:01
Air Blank		0.000	01:02
Subject Sample	# 1	0.217	01:03
Breath Volume		2.359	01:03
Air Blank		0.000	01:03
Air Blank		0.000	01:05
Subject Sample	# 2	0.211	01:06
Breath Volume		3.464	01:06
Air Blank		0.000	01:06
Control Test		0.080	01:07
Air Blank		0.000	01:07
Diagnostics Check		OK	01:07

Data Download Date/Time: 11/27/2017 10:15

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006758

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 10/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 123456789
DRY GAS EXPIRY DATE: 11/11/2018
VIOLATION CODE: DUI
ARREST TIME: 21:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:10

SUBJECT'S LAST NAME: AMADO CRUZ

DATE/TIME OF TEST: 11/25/2017 21:43:45

Diagnostics Check		OK	21:36
Air Blank		0.000	21:37
Control Test		0.081	21:37
Air Blank		0.000	21:38
Subject Sample	# 1	0.193	21:38
Breath Volume		2.011	21:38
Air Blank		0.000	21:39
Air Blank		0.000	21:41
Subject Sample	# 2	0.206	21:42
Breath Volume		3.152	21:42
Air Blank		0.000	21:42
Control Test		0.081	21:43
Air Blank		0.000	21:43
Diagnostics Check		OK	21:43

Data Download Date/Time: 11/27/2017 10:15

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006758

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 10/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 123456789
DRY GAS EXPIRY DATE: 11/11/2018
VIOLATION CODE: DUI
ARREST TIME: 22:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:33

SUBJECT'S LAST NAME: GARCIA LOPEZ

DATE/TIME OF TEST: 11/26/2017 23:17:35

Diagnostics Check		OK	23:11	
Air Blank		0.000	23:12	
Control Test		0.080	23:12	
Air Blank		0.000	23:12	
Subject Sample	# 1	0.221	23:14	
Breath Volume		1.781	23:14	
Air Blank		0.000	23:14	
Air Blank			23:16	RFI Detect
Air Blank		0.000	23:16	
Control Test		0.081	23:16	
Air Blank		0.000	23:17	
Diagnostics Check		OK	23:17	

Data Download Date/Time: 11/27/2017 10:15

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006758

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 10/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 123456789
DRY GAS EXPIRY DATE: 11/11/2018
VIOLATION CODE: DUI
ARREST TIME: 22:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:33

SUBJECT'S LAST NAME: GARCIA LOPEZ

DATE/TIME OF TEST: 11/26/2017 23:28:34

Diagnostics Check		OK	23:21
Air Blank		0.000	23:21
Control Test		0.081	23:22
Air Blank		0.000	23:22
Subject Sample	# 1	0.216	23:23
Breath Volume		2.292	23:23
Air Blank		0.000	23:24
Air Blank		0.000	23:26
Subject Sample	# 2	0.218	23:26
Breath Volume		1.609	23:26
Air Blank		0.000	23:27
Control Test		0.080	23:27
Air Blank		0.000	23:28
Diagnostics Check		OK	23:28

Data Download Date/Time: 11/27/2017 10:15

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006760

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOWMAN,JOANNA,N
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BOWMAN,JOANNA
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 11/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:59

SUBJECT'S LAST NAME: LUNA

DATE/TIME OF TEST: 11/23/2017 04:28:45

Diagnostics Check		OK	04:22
Air Blank		0.000	04:22
Control Test		0.082	04:23
Air Blank		0.000	04:23
Subject Sample	# 1	0.153	04:24
Breath Volume		3.027	04:24
Air Blank		0.000	04:24
Air Blank		0.000	04:26
Subject Sample	# 2	0.146	04:27
Breath Volume		3.085	04:27
Air Blank		0.000	04:27
Control Test		0.081	04:28
Air Blank		0.000	04:28
Diagnostics Check		OK	04:28

Data Download Date/Time: 12/01/2017 22:23

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006765

INSTRUMENT REGISTERED TO: FHP TROOP G
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HEALY,G,B
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: HEALY,G
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 11/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 814181
DRY GAS EXPIRY DATE: 07/04/2020
VIOLATION CODE: DUI
ARREST TIME: 01:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: WHITLOW

DATE/TIME OF TEST: 11/02/2017 01:23:40

Diagnostics Check		OK	01:20	
Air Blank		0.000	01:21	
Control Test		0.080	01:21	
Air Blank		0.000	01:22	
Subject Sample	# 1		01:22	Subject Test Refused
Breath Volume		0.000	01:22	
Air Blank		0.000	01:22	
Control Test		0.079	01:23	
Air Blank		0.000	01:23	
Diagnostics Check		OK	01:23	

Data Download Date/Time: 12/15/2017 21:02

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006765

INSTRUMENT REGISTERED TO: FHP TROOP G
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HEALY,GREGORY,B
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: HEALY,GREGORY
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 11/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:09

SUBJECT'S LAST NAME: TOTH

DATE/TIME OF TEST: 11/04/2017 22:36:09

Diagnostics Check		OK	22:29
Air Blank		0.000	22:30
Control Test		0.079	22:30
Air Blank		0.000	22:30
Subject Sample	# 1	0.147	22:31
Breath Volume		2.035	22:31
Air Blank		0.000	22:32
Air Blank		0.000	22:34
Subject Sample	# 2	0.148	22:34
Breath Volume		3.355	22:34
Air Blank		0.000	22:35
Control Test		0.079	22:35
Air Blank		0.000	22:36
Diagnostics Check		OK	22:36

Data Download Date/Time: 12/15/2017 21:02

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006765

INSTRUMENT REGISTERED TO: FHP TROOP G
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HEALY,GREGORY,B
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: HEALY,GREGORY
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 11/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: RODEN

DATE/TIME OF TEST: 11/19/2017 00:36:36

Diagnostics Check		OK	00:30
Air Blank		0.000	00:30
Control Test		0.079	00:31
Air Blank		0.000	00:31
Subject Sample	# 1	0.122	00:32
Breath Volume		4.226	00:32
Air Blank		0.000	00:32
Air Blank		0.000	00:34
Subject Sample	# 2	0.125	00:35
Breath Volume		4.640	00:35
Air Blank		0.000	00:35
Control Test		0.081	00:35
Air Blank		0.000	00:36
Diagnostics Check		OK	00:36

Data Download Date/Time: 12/15/2017 21:02

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006765

INSTRUMENT REGISTERED TO: FHP TROOP G
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HEALY,GREGORY,B
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: HEALY,GREGORY
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 11/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:39

SUBJECT'S LAST NAME: BASCHON

DATE/TIME OF TEST: 11/25/2017 23:05:38

Diagnostics Check		OK	22:59
Air Blank		0.000	22:59
Control Test		0.079	23:00
Air Blank		0.000	23:00
Subject Sample	# 1	0.111	23:00
Breath Volume		1.562	23:00
Air Blank		0.000	23:01
Air Blank		0.000	23:03
Subject Sample	# 2	0.125	23:04
Breath Volume		1.308	23:04
Air Blank		0.000	23:04
Control Test		0.081	23:05
Air Blank		0.000	23:05
Diagnostics Check		OK	23:05

Data Download Date/Time: 12/15/2017 21:02

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006931

INSTRUMENT REGISTERED TO: FT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERSON,GABRIEL,A
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: ROBERSON,GABRIEL
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 11/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A2
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:12

SUBJECT'S LAST NAME: EDDINS

DATE/TIME OF TEST: 11/23/2017 04:50:44

Diagnostics Check		OK	04:39	
Air Blank		0.000	04:39	
Control Test		0.079	04:40	
Air Blank		0.000	04:40	
Subject Sample	# 1		04:43	No Sample Provided
Breath Volume		0.000	04:43	
Air Blank		0.000	04:44	
Air Blank		0.000	04:46	
Subject Sample	# 2		04:49	No Sample Provided
Breath Volume		0.000	04:49	
Air Blank		0.000	04:49	
Control Test		0.080	04:50	
Air Blank		0.000	04:50	
Diagnostics Check		OK	04:50	

Data Download Date/Time: 11/27/2017 10:53

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006931

INSTRUMENT REGISTERED TO: FT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERSON,GABRIEL,A
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: ROBERSON,GABRIEL
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 11/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A2
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:40

SUBJECT'S LAST NAME: EDDINS

DATE/TIME OF TEST: 11/23/2017 05:09:35

Diagnostics Check		OK	05:03
Air Blank		0.000	05:03
Control Test		0.079	05:04
Air Blank		0.000	05:04
Subject Sample	# 1	0.173	05:05
Breath Volume		1.804	05:05
Air Blank		0.000	05:05
Air Blank		0.000	05:07
Subject Sample	# 2	0.165	05:07
Breath Volume		2.605	05:07
Air Blank		0.000	05:08
Control Test		0.077	05:08
Air Blank		0.000	05:09
Diagnostics Check		OK	05:09

Data Download Date/Time: 11/27/2017 10:53

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.