

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000999

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,RAY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: RIVERA,CARLOS
ARRESTING OFFICER AGENCY: Altamonte Springs PD
LAST AGENCY INSPECTION: 11/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 00:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: WISE JR.

DATE/TIME OF TEST: 11/29/2017 01:27:14

Diagnostics Check	OK	01:20
Air Blank	0.000	01:21
Control Test	0.078	01:21
Air Blank	0.000	01:22
Subject Sample # 1	0.259	01:22
Breath Volume	2.742	01:22
Air Blank	0.000	01:23
Air Blank	0.000	01:25
Subject Sample # 2	0.251	01:25
Breath Volume	3.042	01:25
Air Blank	0.000	01:26
Control Test	0.079	01:26
Air Blank	0.000	01:27
Diagnostics Check	OK	01:27

Data Download Date/Time: 01/02/2018 23:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000999

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,RAY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: KINKOPF,TIMOTHY
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 11/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 03:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: ISENMANN-FOSTER

DATE/TIME OF TEST: 11/29/2017 04:07:53

Diagnostics Check	OK	04:00
Air Blank	0.000	04:01
Control Test	0.079	04:01
Air Blank	0.000	04:02
Subject Sample # 1	0.157	04:02
Breath Volume	2.128	04:02
Air Blank	0.000	04:03
Air Blank	0.000	04:05
Subject Sample # 2	0.146	04:06
Breath Volume	1.230	04:06
Air Blank	0.000	04:07
Control Test	0.078	04:07
Air Blank	0.000	04:07
Diagnostics Check	OK	04:07

Data Download Date/Time: 01/02/2018 23:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000999

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,RAY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: MOREL,ANTHONY
ARRESTING OFFICER AGENCY: Winter Springs PD
LAST AGENCY INSPECTION: 11/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 20:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:45

SUBJECT'S LAST NAME: NOLIN

DATE/TIME OF TEST: 11/29/2017 21:13:35

Diagnostics Check	OK	21:07
Air Blank	0.000	21:07
Control Test	0.079	21:08
Air Blank	0.000	21:08
Subject Sample # 1	0.093	21:09
Breath Volume	1.578	21:09
Air Blank	0.000	21:09
Air Blank	0.000	21:11
Subject Sample # 2	0.096	21:12
Breath Volume	1.832	21:12
Air Blank	0.000	21:12
Control Test	0.079	21:13
Air Blank	0.000	21:13
Diagnostics Check	OK	21:13

Data Download Date/Time: 01/02/2018 23:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000999

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,RAY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: REHDER,PATRICK
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 11/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 22:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: DONALDSON

DATE/TIME OF TEST: 11/29/2017 23:01:47

Diagnostics Check	OK	22:55
Air Blank	0.000	22:55
Control Test	0.078	22:56
Air Blank	0.000	22:56
Subject Sample # 1	0.226	22:57
Breath Volume	1.546	22:57
Air Blank	0.000	22:57
Air Blank	0.000	22:59
Subject Sample # 2	0.240	23:00
Breath Volume	1.859	23:00
Air Blank	0.000	23:00
Control Test	0.079	23:01
Air Blank	0.000	23:01
Diagnostics Check	OK	23:01

Data Download Date/Time: 01/02/2018 23:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001002

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTWRIGHT JR,GARY,W
BREATH TEST OPERATOR AGENCY: Hardee County SO
ARRESTING OFFICER: BOGUSKY,DENNIS
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 10/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 23:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: RAMIREZ-RODRIGUEZ

DATE/TIME OF TEST: 11/13/2017 00:27:08

Diagnostics Check	OK	00:20
Air Blank	0.000	00:20
Control Test	0.079	00:21
Air Blank	0.000	00:21
Subject Sample # 1	0.152	00:22
Breath Volume	3.398	00:22
Air Blank	0.000	00:23
Air Blank	0.000	00:24
Subject Sample # 2	0.152	00:25
Breath Volume	3.437	00:25
Air Blank	0.000	00:26
Control Test	0.079	00:26
Air Blank	0.000	00:27
Diagnostics Check	OK	00:27

Data Download Date/Time: 11/15/2017 08:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001002

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FOWLER,JOHN,T
BREATH TEST OPERATOR AGENCY: Hardee County SO
ARRESTING OFFICER: SPENCER,ROBERT
ARRESTING OFFICER AGENCY: Wauchula PD
LAST AGENCY INSPECTION: 11/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 00:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:28

SUBJECT'S LAST NAME: WRIGHT

DATE/TIME OF TEST: 11/18/2017 00:51:33

Diagnostics Check	OK	00:48	
Air Blank	0.000	00:49	
Control Test	0.079	00:49	
Air Blank	0.000	00:50	
Subject Sample	# 1	00:50	Subject Test Refused
Breath Volume	0.000	00:50	
Air Blank	0.000	00:50	
Control Test	0.080	00:50	
Air Blank	0.000	00:51	
Diagnostics Check	OK	00:51	

Data Download Date/Time: 12/27/2017 16:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001002

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTWRIGHT JR,GARY,W
BREATH TEST OPERATOR AGENCY: Hardee County SO
ARRESTING OFFICER: DAWSON,TYLER
ARRESTING OFFICER AGENCY: Hardee County SO
LAST AGENCY INSPECTION: 11/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:07

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 11/22/2017 03:35:15

Diagnostics Check	OK	03:29
Air Blank	0.000	03:29
Control Test	0.080	03:29
Air Blank	0.000	03:30
Subject Sample # 1	0.151	03:30
Breath Volume	1.269	03:30
Air Blank	0.000	03:31
Air Blank	0.000	03:33
Subject Sample # 2	0.148	03:33
Breath Volume	1.160	03:33
Air Blank	0.000	03:34
Control Test	0.079	03:34
Air Blank	0.000	03:35
Diagnostics Check	OK	03:35

Data Download Date/Time: 12/27/2017 16:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001002

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTWRIGHT JR,GARY,W
BREATH TEST OPERATOR AGENCY: Hardee County SO
ARRESTING OFFICER: HART,LYLE
ARRESTING OFFICER AGENCY: Hardee County SO
LAST AGENCY INSPECTION: 11/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 21:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:21

SUBJECT'S LAST NAME: CRAWFORD

DATE/TIME OF TEST: 11/29/2017 22:56:01

Diagnostics Check	OK	22:50
Air Blank	0.000	22:50
Control Test	0.079	22:50
Air Blank	0.000	22:51
Subject Sample # 1	0.000	22:51
Breath Volume	1.550	22:51
Air Blank	0.000	22:52
Air Blank	0.000	22:54
Subject Sample # 2	0.000	22:54
Breath Volume	2.226	22:54
Air Blank	0.000	22:55
Control Test	0.079	22:55
Air Blank	0.000	22:55
Diagnostics Check	OK	22:56

Data Download Date/Time: 12/27/2017 16:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUARTE,CHRISTOPHER,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: DUARTE,CHRISTOPHER
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 10/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 00:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: LITTLE

DATE/TIME OF TEST: 11/01/2017 02:03:53

Diagnostics Check	OK	01:51	
Air Blank	0.000	01:51	
Control Test	0.081	01:52	
Air Blank	0.000	01:52	
Subject Sample # 1	0.000	01:53	
Breath Volume	1.378	01:53	
Air Blank	0.000	01:53	
Air Blank	0.000	01:56	
Subject Sample # 2	0.000	01:59	Volume Not Met
Breath Volume	0.917	01:59	
Air Blank	0.000	01:59	
Air Blank	0.000	02:01	
Subject Sample # 3	0.000	02:02	
Breath Volume	1.289	02:02	
Air Blank	0.000	02:02	
Control Test	0.080	02:03	
Air Blank	0.000	02:03	
Diagnostics Check	OK	02:03	

Data Download Date/Time: 11/29/2017 22:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORAN,SEAN,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: DENARD,JUSTIN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 07:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:13

SUBJECT'S LAST NAME: JURABE KAUFMEN

DATE/TIME OF TEST: 11/01/2017 09:36:37

Diagnostics Check	OK	09:33	
Air Blank	0.000	09:34	
Control Test	0.081	09:34	
Air Blank	0.000	09:35	
Subject Sample	# 1	09:35	Subject Test Refused
Breath Volume	0.000	09:35	
Air Blank	0.000	09:35	
Control Test	0.082	09:36	
Air Blank	0.000	09:36	
Diagnostics Check	OK	09:36	

Data Download Date/Time: 11/29/2017 22:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KASELAK,STEPHEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 03:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:24

SUBJECT'S LAST NAME: OLIVA

DATE/TIME OF TEST: 11/02/2017 04:12:39

Diagnostics Check	OK	04:06
Air Blank	0.000	04:06
Control Test	0.081	04:06
Air Blank	0.000	04:07
Subject Sample # 1	0.116	04:08
Breath Volume	1.359	04:08
Air Blank	0.000	04:08
Air Blank	0.000	04:10
Subject Sample # 2	0.117	04:11
Breath Volume	1.292	04:11
Air Blank	0.000	04:11
Control Test	0.081	04:12
Air Blank	0.000	04:12
Diagnostics Check	OK	04:12

Data Download Date/Time: 11/29/2017 22:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAIR, RONALD, J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAIR, RONALD
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 20:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:47

SUBJECT'S LAST NAME: MULLIGAN

DATE/TIME OF TEST: 11/03/2017 21:15:28

Diagnostics Check	OK	21:07
Air Blank	0.000	21:07
Control Test	0.081	21:08
Air Blank	0.000	21:08
Subject Sample # 1	0.199	21:10
Breath Volume	1.906	21:10
Air Blank	0.000	21:10
Air Blank	0.000	21:12
Subject Sample # 2	0.184	21:13
Breath Volume	1.671	21:13
Air Blank	0.000	21:14
Control Test	0.080	21:14
Air Blank	0.000	21:15
Diagnostics Check	OK	21:15

Data Download Date/Time: 11/29/2017 22:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KASELAK,STEPHEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 02:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:38

SUBJECT'S LAST NAME: DIGMAN

DATE/TIME OF TEST: 11/04/2017 03:10:00

Diagnostics Check	OK	03:06	
Air Blank	0.000	03:07	
Control Test	0.080	03:07	
Air Blank	0.000	03:08	
Subject Sample	# 1	03:08	Subject Test Refused
Breath Volume	0.000	03:08	
Air Blank	0.000	03:09	
Control Test	0.081	03:09	
Air Blank	0.000	03:09	
Diagnostics Check	OK	03:10	

Data Download Date/Time: 11/29/2017 22:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAIR, RONALD, J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAIR, RONALD
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 21:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:51

SUBJECT'S LAST NAME: SIM

DATE/TIME OF TEST: 11/04/2017 22:19:57

Diagnostics Check	OK	22:13
Air Blank	0.000	22:13
Control Test	0.080	22:14
Air Blank	0.000	22:14
Subject Sample # 1	0.229	22:15
Breath Volume	1.679	22:15
Air Blank	0.000	22:15
Air Blank	0.000	22:17
Subject Sample # 2	0.234	22:18
Breath Volume	1.898	22:18
Air Blank	0.000	22:19
Control Test	0.080	22:19
Air Blank	0.000	22:19
Diagnostics Check	OK	22:19

Data Download Date/Time: 11/29/2017 22:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KASELAK,STEPHEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 22:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:48

SUBJECT'S LAST NAME: LADUCA

DATE/TIME OF TEST: 11/04/2017 23:21:37

Diagnostics Check	OK	23:09	
Air Blank	0.000	23:10	
Control Test	0.081	23:10	
Air Blank	0.000	23:11	
Subject Sample	# 1 0.165	23:14	Volume Not Met
Breath Volume	0.476	23:14	
Air Blank	0.000	23:15	
Air Blank	0.000	23:16	
Subject Sample	# 2 0.166	23:19	Volume Not Met
Breath Volume	0.871	23:19	
Air Blank	0.000	23:20	
Control Test	0.079	23:20	
Air Blank	0.000	23:21	
Diagnostics Check	OK	23:21	

Data Download Date/Time: 11/29/2017 22:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KASELAK,STEPHEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 22:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:48

SUBJECT'S LAST NAME: LADUCA

DATE/TIME OF TEST: 11/04/2017 23:34:27

Diagnostics Check	OK	23:22	
Air Blank	0.000	23:23	
Control Test	0.079	23:23	
Air Blank	0.000	23:24	
Subject Sample	# 1 0.164	23:27	Volume Not Met
Breath Volume	0.800	23:27	
Air Blank	0.000	23:27	
Air Blank	0.000	23:29	
Subject Sample	# 2 0.167	23:32	Volume Not Met
Breath Volume	0.066	23:32	
Air Blank	0.000	23:33	
Control Test	0.079	23:33	
Air Blank	0.000	23:34	
Diagnostics Check	OK	23:34	

Data Download Date/Time: 11/29/2017 22:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REFUSE, RONY,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: REFUSE, RONY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 10/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 21:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:59

SUBJECT'S LAST NAME: MAVRICK

DATE/TIME OF TEST: 11/11/2017 22:31:01

Diagnostics Check	OK	22:19	
Air Blank	0.000	22:20	
Control Test	0.080	22:20	
Air Blank	0.000	22:20	
Subject Sample	# 1 0.000	22:24	Volume Not Met
Breath Volume	0.132	22:24	
Air Blank	0.000	22:24	
Air Blank	0.000	22:26	
Subject Sample	# 2 0.000	22:29	Volume Not Met
Breath Volume	0.398	22:29	
Air Blank	0.000	22:30	
Control Test	0.082	22:30	
Air Blank	0.000	22:30	
Diagnostics Check	OK	22:31	

Data Download Date/Time: 11/29/2017 22:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ANTHONY,M
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: FURTADO

DATE/TIME OF TEST: 11/12/2017 03:46:08

Diagnostics Check	OK	03:39
Air Blank	0.000	03:39
Control Test	0.080	03:40
Air Blank	0.000	03:40
Subject Sample # 1	0.147	03:41
Breath Volume	1.453	03:41
Air Blank	0.000	03:42
Air Blank	0.000	03:44
Subject Sample # 2	0.143	03:44
Breath Volume	1.527	03:44
Air Blank	0.000	03:45
Control Test	0.080	03:45
Air Blank	0.000	03:46
Diagnostics Check	OK	03:46

Data Download Date/Time: 11/29/2017 22:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEXANDER,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ELLIS,ALEXANDER
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 03:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: LENOX

DATE/TIME OF TEST: 11/12/2017 04:30:46

Diagnostics Check	OK	04:26	
Air Blank	0.000	04:27	
Control Test	0.080	04:27	
Air Blank	0.000	04:27	
Subject Sample	# 1	04:29	Subject Test Refused
Breath Volume	0.000	04:29	
Air Blank	0.000	04:29	
Control Test	0.081	04:30	
Air Blank	0.000	04:30	
Diagnostics Check	OK	04:30	

Data Download Date/Time: 11/29/2017 22:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KASELAK,STEPHEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 03:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:48

SUBJECT'S LAST NAME: PIRONE

DATE/TIME OF TEST: 11/12/2017 06:21:25

Diagnostics Check	OK	06:15
Air Blank	0.000	06:15
Control Test	0.081	06:16
Air Blank	0.000	06:16
Subject Sample # 1	0.110	06:17
Breath Volume	1.492	06:17
Air Blank	0.000	06:17
Air Blank	0.000	06:19
Subject Sample # 2	0.106	06:19
Breath Volume	1.558	06:19
Air Blank	0.000	06:20
Control Test	0.080	06:20
Air Blank	0.000	06:21
Diagnostics Check	OK	06:21

Data Download Date/Time: 11/29/2017 22:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZELINSKY,WAYNE,L
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ZELINSKY,WAYNE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/25/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 15:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:10

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 11/12/2017 16:32:49

Diagnostics Check	OK	16:32	
Air Blank		16:32	RFI Detect
Air Blank	0.000	16:32	

Data Download Date/Time: 11/29/2017 22:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZELINSKY,WAYNE,L
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ZELINSKY,WAYNE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 15:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:10

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 11/12/2017 16:35:21

Diagnostics Check	OK	16:34	
Air Blank		16:35	RFI Detect
Air Blank		16:35	RFI Detect

Data Download Date/Time: 11/29/2017 22:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZELINSKY,WAYNE,L
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ZELINSKY,WAYNE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 15:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:10

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 11/12/2017 16:47:11

Diagnostics Check	OK	16:43	
Air Blank	0.000	16:43	
Control Test	0.081	16:43	
Air Blank	0.000	16:44	
Subject Sample	# 1	16:45	Subject Test Refused
Breath Volume	0.070	16:45	
Air Blank	0.000	16:46	
Control Test	0.081	16:46	
Air Blank	0.000	16:47	
Diagnostics Check	OK	16:47	

Data Download Date/Time: 11/29/2017 22:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCKENZIE,DEPUTY,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: SMALLIOS,DEPUTY
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 13:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:55

SUBJECT'S LAST NAME: LAMB

DATE/TIME OF TEST: 11/15/2017 14:19:41

Diagnostics Check	OK	14:16	
Air Blank	0.000	14:16	
Control Test	0.080	14:16	
Air Blank	0.000	14:17	
Subject Sample	# 1	14:18	Subject Test Refused
Breath Volume	0.000	14:18	
Air Blank	0.000	14:18	
Control Test	0.078	14:19	
Air Blank	0.000	14:19	
Diagnostics Check	OK	14:19	

Data Download Date/Time: 11/29/2017 22:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNAVELY,KENT,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HANNA,JANE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 06:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:01

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 11/17/2017 07:24:41

Diagnostics Check	OK	07:21	
Air Blank	0.000	07:22	
Control Test	0.080	07:22	
Air Blank	0.000	07:23	
Subject Sample	# 1	07:23	Subject Test Refused
Breath Volume	0.000	07:23	
Air Blank	0.000	07:23	
Control Test	0.081	07:24	
Air Blank	0.000	07:24	
Diagnostics Check	OK	07:24	

Data Download Date/Time: 11/29/2017 22:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILTSE,DEPUTY W.,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WILTSE,DEPUTY W.
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 00:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:07

SUBJECT'S LAST NAME: OLSON

DATE/TIME OF TEST: 11/18/2017 01:34:17

Diagnostics Check	OK	01:27
Air Blank	0.000	01:27
Control Test	0.081	01:28
Air Blank	0.000	01:28
Subject Sample # 1	0.065	01:29
Breath Volume	1.355	01:29
Air Blank	0.000	01:30
Air Blank	0.000	01:32
Subject Sample # 2	0.068	01:32
Breath Volume	1.511	01:32
Air Blank	0.000	01:33
Control Test	0.080	01:33
Air Blank	0.000	01:34
Diagnostics Check	OK	01:34

Data Download Date/Time: 11/29/2017 22:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PERSAUD,COMPTON,D
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: PERSAUD,COMPTON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 03:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:22

SUBJECT'S LAST NAME: SAMJI

DATE/TIME OF TEST: 11/18/2017 04:51:58

Diagnostics Check	OK	04:43
Air Blank	0.000	04:43
Control Test	0.080	04:43
Air Blank	0.000	04:44
Subject Sample # 1	0.160	04:47
Breath Volume	1.460	04:47
Air Blank	0.000	04:47
Air Blank	0.000	04:49
Subject Sample # 2	0.167	04:50
Breath Volume	1.281	04:50
Air Blank	0.000	04:51
Control Test	0.079	04:51
Air Blank	0.000	04:51
Diagnostics Check	OK	04:51

Data Download Date/Time: 11/29/2017 22:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PERSAUD,COMPTON,D
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: PERSAUD,COMPTON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 01:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:39

SUBJECT'S LAST NAME: MESSANO

DATE/TIME OF TEST: 11/19/2017 02:02:52

Diagnostics Check	OK	01:59	
Air Blank	0.000	01:59	
Control Test	0.080	02:00	
Air Blank	0.000	02:00	
Subject Sample	# 1	02:01	Subject Test Refused
Breath Volume	0.000	02:01	
Air Blank	0.000	02:01	
Control Test	0.081	02:02	
Air Blank	0.000	02:02	
Diagnostics Check	OK	02:02	

Data Download Date/Time: 11/29/2017 22:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCKENZIE,DEPUTY,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: FALLAHEE,DEPUTY
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 12:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:30

SUBJECT'S LAST NAME: KREUSCH

DATE/TIME OF TEST: 11/20/2017 14:01:46

Diagnostics Check	OK	13:50	
Air Blank	0.000	13:50	
Control Test	0.079	13:50	
Air Blank	0.000	13:51	
Subject Sample	# 1 0.218	13:54	Volume Not Met
Breath Volume	0.726	13:54	
Air Blank	0.000	13:55	
Air Blank	0.000	13:57	
Subject Sample	# 2 0.203	14:00	Volume Not Met
Breath Volume	0.683	14:00	
Air Blank	0.000	14:00	
Control Test	0.080	14:01	
Air Blank	0.000	14:01	
Diagnostics Check	OK	14:01	

Data Download Date/Time: 11/29/2017 22:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACOBY,LARRY,K
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: JACOBY,LARRY
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 02:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:22

SUBJECT'S LAST NAME: MEURER

DATE/TIME OF TEST: 11/23/2017 03:54:36

Diagnostics Check	OK	03:43	
Air Blank	0.000	03:43	
Control Test	0.081	03:43	
Air Blank	0.000	03:44	
Subject Sample	# 1 0.107	03:47	Volume Not Met
Breath Volume	0.167	03:47	
Air Blank	0.000	03:48	
Air Blank	0.000	03:49	
Subject Sample	# 2 0.119	03:53	Volume Not Met
Breath Volume	0.156	03:53	
Air Blank	0.000	03:53	
Control Test	0.080	03:54	
Air Blank	0.000	03:54	
Diagnostics Check	OK	03:54	

Data Download Date/Time: 11/29/2017 22:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KASELAK,STEPHEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 22:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:59

SUBJECT'S LAST NAME: BAILEY

DATE/TIME OF TEST: 11/24/2017 23:25:43

Diagnostics Check	OK	23:22	
Air Blank	0.000	23:23	
Control Test	0.080	23:23	
Air Blank	0.000	23:24	
Subject Sample	# 1	23:24	Subject Test Refused
Breath Volume	0.000	23:24	
Air Blank	0.000	23:24	
Control Test	0.080	23:25	
Air Blank	0.000	23:25	
Diagnostics Check	OK	23:25	

Data Download Date/Time: 11/29/2017 22:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH, WILLIAM, T
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SMITH, WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 10/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:43

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 11/25/2017 01:10:09

Diagnostics Check	OK	01:03
Air Blank	0.000	01:03
Control Test	0.080	01:04
Air Blank	0.000	01:04
Subject Sample # 1	0.048	01:05
Breath Volume	1.175	01:05
Air Blank	0.000	01:06
Air Blank	0.000	01:08
Subject Sample # 2	0.046	01:08
Breath Volume	1.187	01:08
Air Blank	0.000	01:09
Control Test	0.079	01:09
Air Blank	0.000	01:10
Diagnostics Check	OK	01:10

Data Download Date/Time: 11/29/2017 22:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LANEY,DAMON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:04

SUBJECT'S LAST NAME: WHELAN-FODOR

DATE/TIME OF TEST: 11/25/2017 02:35:19

Diagnostics Check	OK	02:26
Air Blank	0.000	02:27
Control Test	0.080	02:27
Air Blank	0.000	02:28
Subject Sample # 1	0.160	02:30
Breath Volume	1.328	02:30
Air Blank	0.000	02:31
Air Blank	0.000	02:33
Subject Sample # 2	0.154	02:33
Breath Volume	1.464	02:33
Air Blank	0.000	02:34
Control Test	0.075	02:34
Air Blank	0.000	02:35
Diagnostics Check	OK	02:35

Data Download Date/Time: 11/29/2017 22:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LANEY,DAMON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 04:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: NEWTON

DATE/TIME OF TEST: 11/25/2017 04:57:23

Diagnostics Check	OK	04:50
Air Blank	0.000	04:50
Control Test	0.080	04:51
Air Blank	0.000	04:51
Subject Sample # 1	0.257	04:52
Breath Volume	1.445	04:52
Air Blank	0.000	04:53
Air Blank	0.000	04:55
Subject Sample # 2	0.254	04:55
Breath Volume	1.492	04:55
Air Blank	0.000	04:56
Control Test	0.079	04:56
Air Blank	0.000	04:57
Diagnostics Check	OK	04:57

Data Download Date/Time: 11/29/2017 22:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LANEY,DAMON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 23:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: ADINOLFI

DATE/TIME OF TEST: 11/26/2017 00:30:22

Diagnostics Check	OK	00:27	
Air Blank	0.000	00:28	
Control Test	0.080	00:28	
Air Blank	0.000	00:28	
Subject Sample	# 1	00:29	Subject Test Refused
Breath Volume	0.000	00:29	
Air Blank	0.000	00:29	
Control Test	0.080	00:29	
Air Blank	0.000	00:30	
Diagnostics Check	OK	00:30	

Data Download Date/Time: 11/29/2017 22:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KASELAK,STEPHEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 00:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: BRYANT

DATE/TIME OF TEST: 11/26/2017 01:49:22

Diagnostics Check	OK	01:40	
Air Blank	0.000	01:41	
Control Test	0.080	01:41	
Air Blank	0.000	01:42	
Subject Sample	# 1	01:45	No Sample Provided
Breath Volume	0.000	01:45	
Air Blank	0.000	01:45	
Air Blank	0.000	01:47	
Subject Sample	# 2	01:48	Subject Test Refused
Breath Volume	0.000	01:48	
Air Blank	0.000	01:48	
Control Test	0.080	01:48	
Air Blank	0.000	01:49	
Diagnostics Check	OK	01:49	

Data Download Date/Time: 11/29/2017 22:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERGER,CHARLES,K
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BERGER,CHARLES
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 02:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:24

SUBJECT'S LAST NAME: BORUTA

DATE/TIME OF TEST: 11/26/2017 03:56:31

Diagnostics Check	OK	03:44	
Air Blank	0.000	03:45	
Control Test	0.080	03:45	
Air Blank	0.000	03:46	
Subject Sample	# 1 0.164	03:49	Slope Not Level
Breath Volume	1.253	03:49	
Air Blank	0.000	03:50	
Air Blank	0.000	03:51	
Subject Sample	# 2 0.165	03:54	Volume Not Met
Breath Volume	0.082	03:54	
Air Blank	0.000	03:55	
Control Test	0.079	03:55	
Air Blank	0.000	03:56	
Diagnostics Check	OK	03:56	

Data Download Date/Time: 11/29/2017 22:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERGER,CHARLES,K
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BERGER,CHARLES
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 02:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:24

SUBJECT'S LAST NAME: BORUTA

DATE/TIME OF TEST: 11/26/2017 04:07:56

Diagnostics Check	OK	03:58
Air Blank	0.000	03:59
Control Test	0.079	03:59
Air Blank	0.000	03:59
Subject Sample # 1	0.171	04:02
Breath Volume	1.121	04:02
Air Blank	0.000	04:03
Air Blank	0.000	04:05
Subject Sample # 2	0.165	04:06
Breath Volume	1.230	04:06
Air Blank	0.000	04:07
Control Test	0.078	04:07
Air Blank	0.000	04:07
Diagnostics Check	OK	04:07

Data Download Date/Time: 11/29/2017 22:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH, WILLIAM, T
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SMITH, WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 10/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: ZERO TOL
ARREST TIME: 04:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:13

SUBJECT'S LAST NAME: ARTHUR

DATE/TIME OF TEST: 11/26/2017 05:38:58

Diagnostics Check	OK	05:35	
Air Blank	0.000	05:35	
Control Test	0.080	05:35	
Air Blank	0.000	05:36	
Subject Sample	# 1	05:37	Subject Test Refused
Breath Volume	0.000	05:37	
Air Blank	0.000	05:38	
Control Test	0.079	05:38	
Air Blank	0.000	05:38	
Diagnostics Check	OK	05:38	

Data Download Date/Time: 11/29/2017 22:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRIFFIN,KYLE,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LEIB,MELANIE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 21:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:55

SUBJECT'S LAST NAME: COSTA

DATE/TIME OF TEST: 11/26/2017 22:25:27

Diagnostics Check	OK	22:18
Air Blank	0.000	22:18
Control Test	0.080	22:19
Air Blank	0.000	22:19
Subject Sample # 1	0.000	22:20
Breath Volume	1.117	22:20
Air Blank	0.000	22:21
Air Blank	0.000	22:23
Subject Sample # 2	0.000	22:24
Breath Volume	1.242	22:24
Air Blank	0.000	22:24
Control Test	0.080	22:24
Air Blank	0.000	22:25
Diagnostics Check	OK	22:25

Data Download Date/Time: 11/29/2017 22:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BICKFORD,BRANT,K
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HECK,WILLIAM
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:49

SUBJECT'S LAST NAME: WALKER BARAJAS

DATE/TIME OF TEST: 11/27/2017 03:14:14

Diagnostics Check	OK	03:09	
Air Blank	0.000	03:10	
Control Test	0.080	03:10	
Air Blank	0.000	03:11	
Subject Sample	# 1	03:12	Subject Test Refused
Breath Volume	0.000	03:12	
Air Blank	0.000	03:13	
Control Test	0.080	03:13	
Air Blank	0.000	03:14	
Diagnostics Check	OK	03:14	

Data Download Date/Time: 11/29/2017 22:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOWATT,NATHAN,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MOWATT,NATHAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:39

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 11/28/2017 02:10:51

Diagnostics Check	OK	01:59	
Air Blank	0.000	01:59	
Control Test	0.080	02:00	
Air Blank	0.000	02:00	
Subject Sample	# 1 0.008	02:03	Volume Not Met
Breath Volume	0.847	02:03	
Air Blank	0.000	02:04	
Air Blank	0.000	02:06	
Subject Sample	# 2 0.000	02:09	Volume Not Met
Breath Volume	0.890	02:09	
Air Blank	0.000	02:09	
Control Test	0.081	02:10	
Air Blank	0.000	02:10	
Diagnostics Check	OK	02:10	

Data Download Date/Time: 11/29/2017 22:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOWATT,NATHAN,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MOWATT,NATHAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:39

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 11/28/2017 02:21:28

Diagnostics Check	OK	02:12
Air Blank	0.000	02:12
Control Test	0.081	02:13
Air Blank	0.000	02:13
Subject Sample # 1	0.000	02:15
Breath Volume	1.105	02:15
Air Blank	0.000	02:15
Air Blank	0.000	02:17
Subject Sample # 2	0.000	02:20
Breath Volume	1.136	02:20
Air Blank	0.000	02:20
Control Test	0.080	02:20
Air Blank	0.000	02:21
Diagnostics Check	OK	02:21

Data Download Date/Time: 11/29/2017 22:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LANEY,DAMON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 20:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:22

SUBJECT'S LAST NAME: SCHELL

DATE/TIME OF TEST: 11/29/2017 20:49:10

Diagnostics Check	OK	20:46	
Air Blank	0.000	20:46	
Control Test	0.080	20:46	
Air Blank	0.000	20:47	
Subject Sample	# 1	20:47	Subject Test Refused
Breath Volume	0.000	20:47	
Air Blank	0.000	20:48	
Control Test	0.080	20:48	
Air Blank	0.000	20:49	
Diagnostics Check	OK	20:49	

Data Download Date/Time: 11/29/2017 22:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEXANDER,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ELLIS,A
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 22:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: URSO

DATE/TIME OF TEST: 11/29/2017 23:19:52

Diagnostics Check	OK	23:11	
Air Blank	0.000	23:11	
Control Test	0.080	23:12	
Air Blank	0.000	23:12	
Subject Sample	# 1	23:15	No Sample Provided
Breath Volume	0.000	23:15	
Air Blank	0.000	23:16	
Air Blank	0.000	23:18	
Subject Sample	# 2	23:18	Subject Test Refused
Breath Volume	0.000	23:18	
Air Blank	0.000	23:18	
Control Test	0.080	23:19	
Air Blank	0.000	23:19	
Diagnostics Check	OK	23:19	

Data Download Date/Time: 12/27/2017 21:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLE,S,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HOLE,S
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 04:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:46

SUBJECT'S LAST NAME: TUCKER

DATE/TIME OF TEST: 11/01/2017 05:18:25

Diagnostics Check	OK	05:07	
Air Blank	0.000	05:07	
Control Test	0.082	05:07	
Air Blank	0.000	05:08	
Subject Sample	# 1	05:11	No Sample Provided
Breath Volume	0.000	05:11	
Air Blank	0.000	05:11	
Air Blank	0.000	05:13	
Subject Sample	# 2	05:17	No Sample Provided
Breath Volume	0.000	05:17	
Air Blank	0.000	05:17	
Control Test	0.082	05:17	
Air Blank	0.000	05:18	
Diagnostics Check	OK	05:18	

Data Download Date/Time: 11/29/2017 20:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERGSTROM,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BERGSTROM,STEPHEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 22:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: REED

DATE/TIME OF TEST: 11/03/2017 23:14:02

Diagnostics Check	OK	23:10	
Air Blank	0.000	23:11	
Control Test	0.081	23:11	
Air Blank	0.000	23:11	
Subject Sample	# 1	23:12	Subject Test Refused
Breath Volume	0.000	23:12	
Air Blank	0.000	23:13	
Control Test	0.082	23:13	
Air Blank	0.000	23:13	
Diagnostics Check	OK	23:14	

Data Download Date/Time: 11/29/2017 20:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SZYMANSKI,JACEK,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GRAHAM,SEAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 18:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:10

SUBJECT'S LAST NAME: COLLINS

DATE/TIME OF TEST: 11/06/2017 19:37:34

Diagnostics Check	OK	19:30
Air Blank	0.000	19:31
Control Test	0.081	19:31
Air Blank	0.000	19:31
Subject Sample # 1	0.000	19:32
Breath Volume	3.296	19:32
Air Blank	0.000	19:33
Air Blank	0.000	19:35
Subject Sample # 2	0.000	19:36
Breath Volume	2.265	19:36
Air Blank	0.000	19:36
Control Test	0.081	19:36
Air Blank	0.000	19:37
Diagnostics Check	OK	19:37

Data Download Date/Time: 11/29/2017 20:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLE,S,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HOLE,S
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 18:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:18

SUBJECT'S LAST NAME: KYLE

DATE/TIME OF TEST: 11/09/2017 19:45:44

Diagnostics Check	OK	19:39
Air Blank	0.000	19:39
Control Test	0.082	19:39
Air Blank	0.000	19:40
Subject Sample # 1	0.285	19:40
Breath Volume	1.718	19:40
Air Blank	0.000	19:41
Air Blank	0.000	19:43
Subject Sample # 2	0.277	19:44
Breath Volume	1.906	19:44
Air Blank	0.000	19:44
Control Test	0.081	19:45
Air Blank	0.000	19:45
Diagnostics Check	OK	19:45

Data Download Date/Time: 11/29/2017 20:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHRISTIE,M,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: CHRISTIE,M
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 19:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:20

SUBJECT'S LAST NAME: FERREIRA FERREIRA

DATE/TIME OF TEST: 11/12/2017 20:48:21

Diagnostics Check	OK	20:41
Air Blank	0.000	20:41
Control Test	0.082	20:41
Air Blank	0.000	20:42
Subject Sample # 1	0.210	20:43
Breath Volume	2.714	20:43
Air Blank	0.000	20:43
Air Blank	0.000	20:45
Subject Sample # 2	0.204	20:46
Breath Volume	2.957	20:46
Air Blank	0.000	20:47
Control Test	0.082	20:47
Air Blank	0.000	20:48
Diagnostics Check	OK	20:48

Data Download Date/Time: 11/29/2017 20:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLAIRE,J,L
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GRIFFIN,S
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 19:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:47

SUBJECT'S LAST NAME: BACON

DATE/TIME OF TEST: 11/13/2017 20:18:45

Diagnostics Check	OK	20:08	
Air Blank	0.000	20:09	
Control Test	0.082	20:09	
Air Blank	0.000	20:10	
Subject Sample	# 1	20:13	No Sample Provided
Breath Volume	0.000	20:13	
Air Blank	0.000	20:13	
Air Blank	0.000	20:15	
Subject Sample	# 2	20:17	Subject Test Refused
Breath Volume	0.000	20:17	
Air Blank	0.000	20:17	
Control Test	0.082	20:18	
Air Blank	0.000	20:18	
Diagnostics Check	OK	20:18	

Data Download Date/Time: 11/29/2017 20:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLAIRE,J,L
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LEFFINGWELL,B
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 19:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:32

SUBJECT'S LAST NAME: WYSE

DATE/TIME OF TEST: 11/13/2017 21:04:01

Diagnostics Check	OK	20:52	
Air Blank	0.000	20:52	
Control Test	0.082	20:53	
Air Blank	0.000	20:53	
Subject Sample	# 1 0.102	20:56	Volume Not Met
Breath Volume	0.894	20:56	
Air Blank	0.000	20:57	
Air Blank	0.000	20:59	
Subject Sample	# 2 0.105	21:02	Volume Not Met
Breath Volume	0.105	21:02	
Air Blank	0.000	21:03	
Control Test	0.081	21:03	
Air Blank	0.000	21:03	
Diagnostics Check	OK	21:04	

Data Download Date/Time: 11/29/2017 20:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLAIRE,J,L
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LEFFINGWELL,B
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 19:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:32

SUBJECT'S LAST NAME: WYSE

DATE/TIME OF TEST: 11/13/2017 21:15:44

Diagnostics Check	OK	21:05	
Air Blank	0.000	21:06	
Control Test	0.081	21:06	
Air Blank	0.000	21:07	
Subject Sample	# 1 0.101	21:10	Volume Not Met
Breath Volume	1.058	21:10	
Air Blank	0.000	21:11	
Air Blank	0.000	21:12	
Subject Sample	# 2	21:14	Subject Test Refused
Breath Volume	0.812	21:14	
Air Blank	0.000	21:14	
Control Test	0.080	21:15	
Air Blank	0.000	21:15	
Diagnostics Check	OK	21:15	

Data Download Date/Time: 11/29/2017 20:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLE,S,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: PANIAGUA,M
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 18:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:05

SUBJECT'S LAST NAME: BENTON

DATE/TIME OF TEST: 11/18/2017 19:36:54

Diagnostics Check	OK	19:25	
Air Blank	0.000	19:25	
Control Test	0.082	19:26	
Air Blank	0.000	19:26	
Subject Sample	# 1	19:29	No Sample Provided
Breath Volume	0.000	19:29	
Air Blank	0.000	19:30	
Air Blank	0.000	19:32	
Subject Sample	# 2	19:35	No Sample Provided
Breath Volume	0.000	19:35	
Air Blank	0.000	19:36	
Control Test	0.082	19:36	
Air Blank	0.000	19:36	
Diagnostics Check	OK	19:36	

Data Download Date/Time: 11/29/2017 20:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LANEY,DAMON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 19:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:16

SUBJECT'S LAST NAME: SCURLOCK

DATE/TIME OF TEST: 11/18/2017 20:43:25

Diagnostics Check	OK	20:37
Air Blank	0.000	20:37
Control Test	0.082	20:38
Air Blank	0.000	20:38
Subject Sample # 1	0.000	20:39
Breath Volume	1.378	20:39
Air Blank	0.000	20:39
Air Blank	0.000	20:41
Subject Sample # 2	0.000	20:42
Breath Volume	1.449	20:42
Air Blank	0.000	20:42
Control Test	0.083	20:42
Air Blank	0.000	20:43
Diagnostics Check	OK	20:43

Data Download Date/Time: 11/29/2017 20:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCKENZIE,DEPUTY,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WILLIAMS,DEPUTY N
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 07:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:18

SUBJECT'S LAST NAME: RISCILE

DATE/TIME OF TEST: 11/19/2017 08:42:31

Diagnostics Check	OK	08:38	
Air Blank	0.000	08:39	
Control Test	0.080	08:39	
Air Blank	0.000	08:39	
Subject Sample	# 1	08:41	Subject Test Refused
Breath Volume	0.000	08:41	
Air Blank	0.000	08:41	
Control Test	0.081	08:41	
Air Blank	0.000	08:42	
Diagnostics Check	OK	08:42	

Data Download Date/Time: 11/29/2017 20:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCKENZIE,DEPUTY,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: SIEM,DEPUTY D
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 09:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:48

SUBJECT'S LAST NAME: ADAMS

DATE/TIME OF TEST: 11/19/2017 10:23:50

Diagnostics Check	OK	10:08	
Air Blank	0.000	10:08	
Control Test	0.080	10:08	
Air Blank	0.000	10:09	
Subject Sample	# 1 0.160	10:12	
Breath Volume	1.554	10:12	
Air Blank	0.000	10:12	
Air Blank	0.000	10:14	
Subject Sample	# 2	10:17	No Sample Provided
Breath Volume	0.000	10:17	
Air Blank	0.000	10:18	
Air Blank	0.000	10:20	
Subject Sample	# 3 0.161	10:22	
Breath Volume	1.179	10:22	
Air Blank	0.000	10:22	
Control Test	0.081	10:23	
Air Blank	0.000	10:23	
Diagnostics Check	OK	10:23	

Data Download Date/Time: 11/29/2017 20:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLAIRE,J,L
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: RANISZEWSKI,A
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 17:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:43

SUBJECT'S LAST NAME: RENTZ

DATE/TIME OF TEST: 11/19/2017 19:18:44

Diagnostics Check	OK	19:04	
Air Blank	0.000	19:04	
Control Test	0.080	19:05	
Air Blank	0.000	19:05	
Subject Sample	# 1 0.337	19:07	
Breath Volume	1.167	19:07	
Air Blank	0.000	19:08	
Air Blank	0.000	19:09	
Subject Sample	# 2 0.353	19:13	Volume Not Met
Breath Volume	0.960	19:13	
Air Blank	0.000	19:14	
Air Blank	0.000	19:15	
Subject Sample	# 3 0.337	19:16	
Breath Volume	2.000	19:16	
Air Blank	0.000	19:17	
Control Test	0.075	19:18	
Air Blank	0.000	19:18	
Diagnostics Check	OK	19:18	

Data Download Date/Time: 11/29/2017 20:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRIFFIN,KYLE,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GRIFFIN,KYLE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 02:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:58

SUBJECT'S LAST NAME: CHATHAM

DATE/TIME OF TEST: 11/27/2017 03:28:07

Diagnostics Check	OK	03:19
Air Blank	0.000	03:20
Control Test	0.080	03:20
Air Blank	0.000	03:21
Subject Sample # 1	0.170	03:22
Breath Volume	1.527	03:22
Air Blank	0.000	03:22
Air Blank	0.000	03:24
Subject Sample # 2	0.174	03:26
Breath Volume	1.507	03:26
Air Blank	0.000	03:27
Control Test	0.077	03:27
Air Blank	0.000	03:28
Diagnostics Check	OK	03:28

Data Download Date/Time: 11/29/2017 20:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH, WILLIAM, T
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SMITH, WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 21:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:04

SUBJECT'S LAST NAME: CANNON

DATE/TIME OF TEST: 11/29/2017 22:32:24

Diagnostics Check	OK	22:26
Air Blank	0.000	22:26
Control Test	0.080	22:26
Air Blank	0.000	22:27
Subject Sample # 1	0.000	22:28
Breath Volume	2.644	22:28
Air Blank	0.000	22:28
Air Blank	0.000	22:30
Subject Sample # 2	0.000	22:31
Breath Volume	3.152	22:31
Air Blank	0.000	22:31
Control Test	0.080	22:31
Air Blank	0.000	22:32
Diagnostics Check	OK	22:32

Data Download Date/Time: 12/27/2017 20:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH, WILLIAM, T
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: REFUSE, RONY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 23:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:09

SUBJECT'S LAST NAME: GANNON

DATE/TIME OF TEST: 11/30/2017 00:46:54

Diagnostics Check	OK	00:39
Air Blank	0.000	00:39
Control Test	0.080	00:40
Air Blank	0.000	00:40
Subject Sample # 1	0.178	00:41
Breath Volume	1.882	00:41
Air Blank	0.000	00:42
Air Blank	0.000	00:44
Subject Sample # 2	0.188	00:45
Breath Volume	1.187	00:45
Air Blank	0.000	00:45
Control Test	0.080	00:46
Air Blank	0.000	00:46
Diagnostics Check	OK	00:46

Data Download Date/Time: 12/27/2017 20:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAGER,KEVIN,M
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: NEWELL,GREG
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 10/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 684131
DRY GAS EXPIRY DATE: 01/23/2020
VIOLATION CODE: DUI
ARREST TIME: 16:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:33

SUBJECT'S LAST NAME: SAN ANDRES

DATE/TIME OF TEST: 11/07/2017 17:01:06

Diagnostics Check	OK	16:53
Air Blank	0.000	16:54
Control Test	0.082	16:54
Air Blank	0.000	16:55
Subject Sample # 1	0.198	16:56
Breath Volume	1.917	16:56
Air Blank	0.000	16:57
Air Blank	0.000	16:58
Subject Sample # 2	0.205	16:59
Breath Volume	1.496	16:59
Air Blank	0.000	17:00
Control Test	0.081	17:00
Air Blank	0.000	17:01
Diagnostics Check	OK	17:01

Data Download Date/Time: 11/30/2017 12:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHEU,STEPHEN,G
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: TURNER,LATORIAN
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 10/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 684131
DRY GAS EXPIRY DATE: 01/23/2020
VIOLATION CODE: DUI
ARREST TIME: 04:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:15

SUBJECT'S LAST NAME: SCHEU

DATE/TIME OF TEST: 11/18/2017 05:41:23

Diagnostics Check	OK	05:38	
Air Blank	0.000	05:39	
Control Test	0.082	05:39	
Air Blank	0.000	05:39	
Subject Sample	# 1	05:40	Subject Test Refused
Breath Volume	0.000	05:40	
Air Blank	0.000	05:40	
Control Test	0.082	05:40	
Air Blank	0.000	05:41	
Diagnostics Check	OK	05:41	

Data Download Date/Time: 11/30/2017 12:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACKSON,CHRISTOPHER,M
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: JACKSON,CHRISTOPHER
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 10/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 684131
DRY GAS EXPIRY DATE: 01/23/2020
VIOLATION CODE: DUI
ARREST TIME: 15:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:50

SUBJECT'S LAST NAME: CROW-HENDRICK

DATE/TIME OF TEST: 11/18/2017 16:43:49

Diagnostics Check	OK	16:37
Air Blank	0.000	16:38
Control Test	0.082	16:38
Air Blank	0.000	16:38
Subject Sample # 1	0.096	16:39
Breath Volume	1.367	16:39
Air Blank	0.000	16:40
Air Blank	0.000	16:41
Subject Sample # 2	0.082	16:42
Breath Volume	2.046	16:42
Air Blank	0.000	16:42
Control Test	0.079	16:43
Air Blank	0.000	16:43
Diagnostics Check	OK	16:43

Data Download Date/Time: 11/30/2017 12:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,CHAD,B
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: SMITH,CHAD
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 10/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 684131
DRY GAS EXPIRY DATE: 01/23/2020
VIOLATION CODE: DUI
ARREST TIME: 20:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: BRENNAN

DATE/TIME OF TEST: 11/21/2017 21:27:31

Diagnostics Check	OK	21:21
Air Blank	0.000	21:21
Control Test	0.082	21:22
Air Blank	0.000	21:22
Subject Sample # 1	0.208	21:23
Breath Volume	4.007	21:23
Air Blank	0.000	21:23
Air Blank	0.000	21:25
Subject Sample # 2	0.203	21:26
Breath Volume	3.820	21:26
Air Blank	0.000	21:26
Control Test	0.082	21:26
Air Blank	0.000	21:27
Diagnostics Check	OK	21:27

Data Download Date/Time: 11/30/2017 12:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAGER,KEVIN,M
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: RAMIREZ,JOSE
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 10/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 684131
DRY GAS EXPIRY DATE: 01/23/2020
VIOLATION CODE: DUI
ARREST TIME: 15:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:17

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 11/26/2017 15:51:36

Diagnostics Check	OK	15:44
Air Blank	0.000	15:44
Control Test	0.082	15:45
Air Blank	0.000	15:45
Subject Sample # 1	0.248	15:46
Breath Volume	2.363	15:46
Air Blank	0.000	15:47
Air Blank	0.000	15:48
Subject Sample # 2	0.267	15:49
Breath Volume	3.789	15:49
Air Blank	0.000	15:50
Control Test	0.081	15:50
Air Blank	0.000	15:51
Diagnostics Check	OK	15:51

Data Download Date/Time: 11/30/2017 12:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OBYRNE,JAMES,W
BREATH TEST OPERATOR AGENCY: West Melbourne PD
ARRESTING OFFICER: OBYRNE,JAMES
ARRESTING OFFICER AGENCY: West Melbourne PD
LAST AGENCY INSPECTION: 10/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04216080A3
DRY GAS EXPIRY DATE: 03/05/2018
VIOLATION CODE: DUI
ARREST TIME: 15:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:35

SUBJECT'S LAST NAME: STONE

DATE/TIME OF TEST: 11/18/2017 16:05:42

Diagnostics Check	OK	15:59
Air Blank	0.000	16:00
Control Test	0.080	16:00
Air Blank	0.000	16:00
Subject Sample # 1	0.000	16:01
Breath Volume	2.183	16:01
Air Blank	0.000	16:01
Air Blank	0.000	16:03
Subject Sample # 2	0.000	16:04
Breath Volume	1.597	16:04
Air Blank	0.000	16:04
Control Test	0.079	16:05
Air Blank	0.000	16:05
Diagnostics Check	OK	16:05

Data Download Date/Time: 11/27/2017 10:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ABU SOUD,AHMAD,A
BREATH TEST OPERATOR AGENCY: West Melbourne PD
ARRESTING OFFICER: MATHIS,JACOB'
ARRESTING OFFICER AGENCY: West Melbourne PD
LAST AGENCY INSPECTION: 10/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04216080A3
DRY GAS EXPIRY DATE: 03/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:19

SUBJECT'S LAST NAME: MCKENNA

DATE/TIME OF TEST: 11/23/2017 02:47:28

Diagnostics Check	OK	02:40
Air Blank	0.000	02:40
Control Test	0.079	02:41
Air Blank	0.000	02:41
Subject Sample # 1	0.070	02:43
Breath Volume	1.421	02:43
Air Blank	0.000	02:43
Air Blank	0.000	02:45
Subject Sample # 2	0.077	02:45
Breath Volume	1.917	02:45
Air Blank	0.000	02:46
Control Test	0.078	02:46
Air Blank	0.000	02:47
Diagnostics Check	OK	02:47

Data Download Date/Time: 11/27/2017 10:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001010

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOORE,KIMBERLY,N
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: MOORE,KIMBERLY
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 10/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: SANTIAGO

DATE/TIME OF TEST: 11/04/2017 03:14:21

Diagnostics Check	OK	03:07
Air Blank	0.000	03:08
Control Test	0.079	03:08
Air Blank	0.000	03:09
Subject Sample # 1	0.132	03:09
Breath Volume	2.550	03:09
Air Blank	0.000	03:10
Air Blank	0.000	03:12
Subject Sample # 2	0.130	03:12
Breath Volume	2.562	03:12
Air Blank	0.000	03:13
Control Test	0.078	03:13
Air Blank	0.000	03:14
Diagnostics Check	OK	03:14

Data Download Date/Time: 11/16/2017 18:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001010

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROZYCKI,RICHARD,J
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: ALEPRETE,JOSHUA
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 10/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: FABER

DATE/TIME OF TEST: 11/08/2017 00:57:27

Diagnostics Check	OK	00:50
Air Blank	0.000	00:51
Control Test	0.079	00:51
Air Blank	0.000	00:52
Subject Sample # 1	0.265	00:52
Breath Volume	2.781	00:52
Air Blank	0.000	00:53
Air Blank	0.000	00:55
Subject Sample # 2	0.263	00:55
Breath Volume	2.171	00:55
Air Blank	0.000	00:56
Control Test	0.077	00:56
Air Blank	0.000	00:57
Diagnostics Check	OK	00:57

Data Download Date/Time: 11/16/2017 18:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001010

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOORE,KIMBERLY,N
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: MOORE,KIMBERLY
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 10/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:49

SUBJECT'S LAST NAME: REED

DATE/TIME OF TEST: 11/09/2017 20:21:25

Diagnostics Check	OK	20:14
Air Blank	0.000	20:14
Control Test	0.079	20:15
Air Blank	0.000	20:15
Subject Sample # 1	0.199	20:16
Breath Volume	1.785	20:16
Air Blank	0.000	20:17
Air Blank	0.000	20:19
Subject Sample # 2	0.200	20:19
Breath Volume	1.476	20:19
Air Blank	0.000	20:20
Control Test	0.078	20:20
Air Blank	0.000	20:21
Diagnostics Check	OK	20:21

Data Download Date/Time: 11/16/2017 18:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001010

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOORE,KIMBERLY,N
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: MOORE,KIMBERLY
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 10/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 11/10/2017 02:39:17

Diagnostics Check	OK	02:32
Air Blank	0.000	02:33
Control Test	0.079	02:33
Air Blank	0.000	02:33
Subject Sample # 1	0.114	02:34
Breath Volume	1.546	02:34
Air Blank	0.000	02:35
Air Blank	0.000	02:37
Subject Sample # 2	0.111	02:37
Breath Volume	1.585	02:37
Air Blank	0.000	02:38
Control Test	0.079	02:38
Air Blank	0.000	02:39
Diagnostics Check	OK	02:39

Data Download Date/Time: 11/16/2017 18:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001010

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIVINS, HEATHER, C
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: HANIGAN, PATRICK
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 10/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: HODGE

DATE/TIME OF TEST: 11/11/2017 00:21:47

Diagnostics Check	OK	00:15
Air Blank	0.000	00:15
Control Test	0.079	00:15
Air Blank	0.000	00:16
Subject Sample # 1	0.216	00:17
Breath Volume	3.062	00:17
Air Blank	0.000	00:17
Air Blank	0.000	00:19
Subject Sample # 2	0.219	00:20
Breath Volume	3.132	00:20
Air Blank	0.000	00:20
Control Test	0.078	00:21
Air Blank	0.000	00:21
Diagnostics Check	OK	00:21

Data Download Date/Time: 11/16/2017 18:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001010

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOORE,KIMBERLY,N
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: WEIMER,ADAM
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 10/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: FARMER

DATE/TIME OF TEST: 11/11/2017 02:17:24

Diagnostics Check	OK	02:11
Air Blank	0.000	02:11
Control Test	0.078	02:11
Air Blank	0.000	02:12
Subject Sample # 1	0.065	02:12
Breath Volume	2.371	02:12
Air Blank	0.000	02:13
Air Blank	0.000	02:15
Subject Sample # 2	0.062	02:15
Breath Volume	2.507	02:15
Air Blank	0.000	02:16
Control Test	0.078	02:16
Air Blank	0.000	02:17
Diagnostics Check	OK	02:17

Data Download Date/Time: 11/16/2017 18:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSEPILER,JONATHAN,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CARLSON,P
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 10/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:14

SUBJECT'S LAST NAME: MERRYMAN

DATE/TIME OF TEST: 11/01/2017 01:47:00

Diagnostics Check	OK	01:35	
Air Blank	0.000	01:35	
Control Test	0.079	01:35	
Air Blank	0.000	01:36	
Subject Sample	# 1 0.193	01:39	
Breath Volume	2.617	01:39	
Air Blank	0.000	01:40	
Air Blank	0.000	01:42	
Subject Sample	# 2 0.145	01:45	Volume Not Met
Breath Volume	0.601	01:45	
Air Blank		01:46	Purge Fail
Air Blank	0.000	01:47	

Data Download Date/Time: 11/01/2017 12:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON,BRADEN,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CARLSON,P
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 11/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:02

SUBJECT'S LAST NAME: MORRIS

DATE/TIME OF TEST: 11/02/2017 23:27:22

Diagnostics Check	OK	23:24	
Air Blank	0.000	23:24	
Control Test	0.078	23:25	
Air Blank	0.000	23:25	
Subject Sample	# 1	23:25	Subject Test Refused
Breath Volume	0.000	23:25	
Air Blank	0.000	23:26	
Control Test	0.079	23:26	
Air Blank	0.000	23:27	
Diagnostics Check	OK	23:27	

Data Download Date/Time: 11/17/2017 14:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON, BRADEN, J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FILBERT, E
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 11/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:39

SUBJECT'S LAST NAME: PARRA

DATE/TIME OF TEST: 11/03/2017 03:06:30

Diagnostics Check	OK	03:00
Air Blank	0.000	03:00
Control Test	0.079	03:00
Air Blank	0.000	03:01
Subject Sample # 1	0.073	03:01
Breath Volume	1.207	03:01
Air Blank	0.000	03:02
Air Blank	0.000	03:04
Subject Sample # 2	0.072	03:04
Breath Volume	2.812	03:04
Air Blank	0.000	03:05
Control Test	0.079	03:05
Air Blank	0.000	03:06
Diagnostics Check	OK	03:06

Data Download Date/Time: 11/17/2017 14:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON, BRADEN, J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LANE, C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:19

SUBJECT'S LAST NAME: ZARATE PACHECO

DATE/TIME OF TEST: 11/03/2017 04:50:53

Diagnostics Check	OK	04:44
Air Blank	0.000	04:44
Control Test	0.078	04:44
Air Blank	0.000	04:45
Subject Sample # 1	0.100	04:46
Breath Volume	3.156	04:46
Air Blank	0.000	04:46
Air Blank	0.000	04:48
Subject Sample # 2	0.101	04:49
Breath Volume	3.066	04:49
Air Blank	0.000	04:49
Control Test	0.077	04:50
Air Blank	0.000	04:50
Diagnostics Check	OK	04:50

Data Download Date/Time: 11/17/2017 14:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSEPILER,JONATHAN,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CARLSON,P
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 11/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:09

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 11/04/2017 00:35:42

Diagnostics Check	OK	00:29
Air Blank	0.000	00:30
Control Test	0.079	00:30
Air Blank	0.000	00:30
Subject Sample # 1	0.000	00:31
Breath Volume	2.453	00:31
Air Blank	0.000	00:31
Air Blank	0.000	00:33
Subject Sample # 2	0.000	00:34
Breath Volume	2.921	00:34
Air Blank	0.000	00:34
Control Test	0.078	00:35
Air Blank	0.000	00:35
Diagnostics Check	OK	00:35

Data Download Date/Time: 11/17/2017 14:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSEPILER,JONATHAN,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHITE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:36

SUBJECT'S LAST NAME: STUMP

DATE/TIME OF TEST: 11/04/2017 02:03:02

Diagnostics Check	OK	01:56
Air Blank	0.000	01:56
Control Test	0.078	01:57
Air Blank	0.000	01:57
Subject Sample # 1	0.184	01:58
Breath Volume	1.613	01:58
Air Blank	0.000	01:59
Air Blank	0.000	02:00
Subject Sample # 2	0.179	02:01
Breath Volume	1.871	02:01
Air Blank	0.000	02:02
Control Test	0.077	02:02
Air Blank	0.000	02:02
Diagnostics Check	OK	02:03

Data Download Date/Time: 11/17/2017 14:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSEPILER,JONATHAN,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SCOTT,T
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 11/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:47

SUBJECT'S LAST NAME: WYNN

DATE/TIME OF TEST: 11/04/2017 04:11:08

Diagnostics Check	OK	04:08	
Air Blank	0.000	04:08	
Control Test	0.079	04:08	
Air Blank	0.000	04:09	
Subject Sample	# 1	04:09	Subject Test Refused
Breath Volume	0.000	04:09	
Air Blank	0.000	04:10	
Control Test	0.078	04:10	
Air Blank	0.000	04:11	
Diagnostics Check	OK	04:11	

Data Download Date/Time: 11/17/2017 14:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSEPILER,JONATHAN,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: EASTERLY,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:14

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 11/05/2017 00:41:03

Diagnostics Check	OK	00:34
Air Blank	0.000	00:35
Control Test	0.079	00:35
Air Blank	0.000	00:36
Subject Sample # 1	0.100	00:36
Breath Volume	2.710	00:36
Air Blank	0.000	00:37
Air Blank	0.000	00:38
Subject Sample # 2	0.101	00:39
Breath Volume	3.527	00:39
Air Blank	0.000	00:40
Control Test	0.079	00:40
Air Blank	0.000	00:40
Diagnostics Check	OK	00:41

Data Download Date/Time: 11/17/2017 14:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSEPILER,JONATHAN,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: STANLEY,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:28

SUBJECT'S LAST NAME: LEVY

DATE/TIME OF TEST: 11/05/2017 02:01:20

Diagnostics Check	OK	01:49	
Air Blank	0.000	01:49	
Control Test	0.079	01:49	
Air Blank	0.000	01:50	
Subject Sample	# 1 0.000	01:53	Volume Not Met
Breath Volume	0.257	01:53	
Air Blank	0.000	01:53	
Air Blank	0.000	01:55	
Subject Sample	# 2 0.000	01:56	
Breath Volume	1.148	01:56	
Air Blank	0.000	01:56	
Air Blank	0.000	01:58	
Subject Sample	# 3 0.000	01:59	
Breath Volume	1.265	01:59	
Air Blank	0.000	02:00	
Control Test	0.078	02:00	
Air Blank	0.000	02:01	
Diagnostics Check	OK	02:01	

Data Download Date/Time: 11/17/2017 14:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSEPILER,JONATHAN,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: STANLEY,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:26

SUBJECT'S LAST NAME: LEE

DATE/TIME OF TEST: 11/06/2017 00:50:20

Diagnostics Check	OK	00:47	
Air Blank	0.000	00:47	
Control Test	0.078	00:48	
Air Blank	0.000	00:48	
Subject Sample	# 1	00:48	Subject Test Refused
Breath Volume	0.000	00:48	
Air Blank	0.000	00:49	
Control Test	0.078	00:49	
Air Blank	0.000	00:50	
Diagnostics Check	OK	00:50	

Data Download Date/Time: 11/17/2017 14:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSEPILER,JONATHAN,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CAMACHO,K
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 11/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:26

SUBJECT'S LAST NAME: SEXTON

DATE/TIME OF TEST: 11/06/2017 01:49:56

Diagnostics Check	OK	01:47	
Air Blank	0.000	01:47	
Control Test	0.078	01:47	
Air Blank	0.000	01:48	
Subject Sample	# 1	01:48	Subject Test Refused
Breath Volume	0.000	01:48	
Air Blank	0.000	01:48	
Control Test	0.078	01:49	
Air Blank	0.000	01:49	
Diagnostics Check	OK	01:49	

Data Download Date/Time: 11/17/2017 14:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSEPILER,JONATHAN,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: THOMPSON,J
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 11/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:44

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 11/06/2017 05:08:10

Diagnostics Check	OK	05:05	
Air Blank	0.000	05:05	
Control Test	0.079	05:06	
Air Blank	0.000	05:06	
Subject Sample	# 1	05:06	Subject Test Refused
Breath Volume	0.000	05:06	
Air Blank	0.000	05:07	
Control Test	0.078	05:07	
Air Blank	0.000	05:08	
Diagnostics Check	OK	05:08	

Data Download Date/Time: 11/17/2017 14:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON, BRADEN, J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BRYANT, R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:01

SUBJECT'S LAST NAME: JEFFORDS

DATE/TIME OF TEST: 11/17/2017 03:33:37

Diagnostics Check	OK	03:30	
Air Blank	0.000	03:30	
Control Test	0.079	03:31	
Air Blank	0.000	03:31	
Subject Sample	# 1	03:32	Subject Test Refused
Breath Volume	0.000	03:32	
Air Blank	0.000	03:32	
Control Test	0.080	03:32	
Air Blank	0.000	03:33	
Diagnostics Check	OK	03:33	

Data Download Date/Time: 11/17/2017 14:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON,BRADEN,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MATHIS,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:16

SUBJECT'S LAST NAME: POZO

DATE/TIME OF TEST: 11/17/2017 05:40:05

Diagnostics Check	OK	05:37	
Air Blank	0.000	05:37	
Control Test	0.079	05:37	
Air Blank	0.000	05:38	
Subject Sample	# 1	05:38	Subject Test Refused
Breath Volume	0.000	05:38	
Air Blank	0.000	05:39	
Control Test	0.079	05:39	
Air Blank	0.000	05:39	
Diagnostics Check	OK	05:40	

Data Download Date/Time: 11/17/2017 14:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON, BRADEN, J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BROWN, T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:11

SUBJECT'S LAST NAME: MARVIN

DATE/TIME OF TEST: 11/18/2017 00:39:31

Diagnostics Check	OK	00:31
Air Blank	0.000	00:31
Control Test	0.079	00:31
Air Blank	0.000	00:32
Subject Sample # 1	0.154	00:34
Breath Volume	1.218	00:34
Air Blank	0.000	00:35
Air Blank	0.000	00:36
Subject Sample # 2	0.149	00:37
Breath Volume	1.546	00:37
Air Blank	0.000	00:38
Control Test	0.076	00:38
Air Blank	0.000	00:39
Diagnostics Check	OK	00:39

Data Download Date/Time: 12/11/2017 15:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON, BRADEN, J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CAMACHO, K
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:47

SUBJECT'S LAST NAME: ALLRED

DATE/TIME OF TEST: 11/18/2017 03:10:15

Diagnostics Check	OK	03:07	
Air Blank	0.000	03:07	
Control Test	0.078	03:08	
Air Blank	0.000	03:08	
Subject Sample	# 1	03:08	Subject Test Refused
Breath Volume	0.000	03:08	
Air Blank	0.000	03:09	
Control Test	0.078	03:09	
Air Blank	0.000	03:10	
Diagnostics Check	OK	03:10	

Data Download Date/Time: 12/11/2017 15:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON, BRADEN, J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHITE, H
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:02

SUBJECT'S LAST NAME: WALKER

DATE/TIME OF TEST: 11/18/2017 05:28:50

Diagnostics Check	OK	05:22
Air Blank	0.000	05:22
Control Test	0.078	05:23
Air Blank	0.000	05:23
Subject Sample # 1	0.071	05:24
Breath Volume	1.570	05:24
Air Blank	0.000	05:24
Air Blank	0.000	05:26
Subject Sample # 2	0.071	05:27
Breath Volume	1.722	05:27
Air Blank	0.000	05:27
Control Test	0.079	05:28
Air Blank	0.000	05:28
Diagnostics Check	OK	05:28

Data Download Date/Time: 12/11/2017 15:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLARIDA,K,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MILLER,C
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 16:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:05

SUBJECT'S LAST NAME: KENDRICK

DATE/TIME OF TEST: 11/18/2017 18:32:36

Diagnostics Check	OK	18:25
Air Blank	0.000	18:26
Control Test	0.079	18:26
Air Blank	0.000	18:27
Subject Sample # 1	0.264	18:27
Breath Volume	3.000	18:27
Air Blank	0.000	18:28
Air Blank	0.000	18:30
Subject Sample # 2	0.279	18:30
Breath Volume	3.132	18:30
Air Blank	0.000	18:31
Control Test	0.078	18:31
Air Blank	0.000	18:32
Diagnostics Check	OK	18:32

Data Download Date/Time: 12/11/2017 15:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLARIDA,K,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FILBERT,E
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 17:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:25

SUBJECT'S LAST NAME: BROWARD

DATE/TIME OF TEST: 11/18/2017 20:52:29

Diagnostics Check	OK	20:45	
Air Blank	0.000	20:46	
Control Test	0.079	20:46	
Air Blank	0.000	20:47	
Subject Sample	# 1 0.119	20:47	
Breath Volume	1.796	20:47	
Air Blank	0.000	20:48	
Air Blank	0.000	20:50	
Subject Sample	# 2 0.117	20:50	
Breath Volume	3.449	20:50	
Air Blank	0.000	20:51	
Control Test	0.078	20:51	
Air Blank		20:52	RFI Detect
Air Blank	0.000	20:52	

Data Download Date/Time: 12/11/2017 15:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLARIDA,K,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FILBERT,E
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 17:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:53

SUBJECT'S LAST NAME: BROWARD

DATE/TIME OF TEST: 11/18/2017 21:20:45

Diagnostics Check	OK	21:14
Air Blank	0.000	21:14
Control Test	0.079	21:14
Air Blank	0.000	21:15
Subject Sample # 1	0.110	21:15
Breath Volume	2.734	21:15
Air Blank	0.000	21:16
Air Blank	0.000	21:18
Subject Sample # 2	0.107	21:19
Breath Volume	2.980	21:19
Air Blank	0.000	21:19
Control Test	0.078	21:20
Air Blank	0.000	21:20
Diagnostics Check	OK	21:20

Data Download Date/Time: 12/11/2017 15:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON, BRADEN, J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BYRNES, J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:19

SUBJECT'S LAST NAME: SANDERSON

DATE/TIME OF TEST: 11/19/2017 01:46:04

Diagnostics Check	OK	01:39
Air Blank	0.000	01:39
Control Test	0.079	01:40
Air Blank	0.000	01:40
Subject Sample # 1	0.260	01:41
Breath Volume	1.132	01:41
Air Blank	0.000	01:42
Air Blank	0.000	01:43
Subject Sample # 2	0.273	01:44
Breath Volume	2.500	01:44
Air Blank	0.000	01:45
Control Test	0.078	01:45
Air Blank	0.000	01:45
Diagnostics Check	OK	01:46

Data Download Date/Time: 12/11/2017 15:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON, BRADEN, J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: EANES, J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: WARNER

DATE/TIME OF TEST: 11/19/2017 04:22:49

Diagnostics Check	OK	04:15
Air Blank	0.000	04:16
Control Test	0.078	04:16
Air Blank	0.000	04:17
Subject Sample # 1	0.163	04:17
Breath Volume	2.644	04:17
Air Blank	0.000	04:18
Air Blank	0.000	04:20
Subject Sample # 2	0.164	04:21
Breath Volume	2.281	04:21
Air Blank	0.000	04:21
Control Test	0.078	04:22
Air Blank	0.000	04:22
Diagnostics Check	OK	04:22

Data Download Date/Time: 12/11/2017 15:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON, BRADEN, J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HYSMITH, B
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:35

SUBJECT'S LAST NAME: MIRASHI

DATE/TIME OF TEST: 11/20/2017 06:04:03

Diagnostics Check	OK	05:56
Air Blank	0.000	05:56
Control Test	0.080	05:56
Air Blank	0.000	05:57
Subject Sample # 1	0.020	05:59
Breath Volume	1.335	05:59
Air Blank	0.000	05:59
Air Blank	0.000	06:01
Subject Sample # 2	0.019	06:02
Breath Volume	2.183	06:02
Air Blank	0.000	06:03
Control Test	0.080	06:03
Air Blank	0.000	06:03
Diagnostics Check	OK	06:04

Data Download Date/Time: 12/11/2017 15:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURNS,D,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: EANES,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: PERRY

DATE/TIME OF TEST: 11/21/2017 03:57:19

Diagnostics Check	OK	03:50
Air Blank	0.000	03:51
Control Test	0.080	03:51
Air Blank	0.000	03:52
Subject Sample # 1	0.237	03:52
Breath Volume	3.007	03:52
Air Blank	0.000	03:53
Air Blank	0.000	03:55
Subject Sample # 2	0.219	03:55
Breath Volume	2.683	03:55
Air Blank	0.000	03:56
Control Test	0.077	03:56
Air Blank	0.000	03:57
Diagnostics Check	OK	03:57

Data Download Date/Time: 12/11/2017 15:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON, BRADEN, J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ROSSIGNOL, H
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:45

SUBJECT'S LAST NAME: LATIMER

DATE/TIME OF TEST: 11/21/2017 05:12:34

Diagnostics Check	OK	05:05
Air Blank	0.000	05:06
Control Test	0.079	05:06
Air Blank	0.000	05:07
Subject Sample # 1	0.019	05:07
Breath Volume	1.738	05:07
Air Blank	0.000	05:08
Air Blank	0.000	05:10
Subject Sample # 2	0.019	05:10
Breath Volume	1.886	05:10
Air Blank	0.000	05:11
Control Test	0.078	05:11
Air Blank	0.000	05:12
Diagnostics Check	OK	05:12

Data Download Date/Time: 12/11/2017 15:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHITE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:29

SUBJECT'S LAST NAME: MACDONALD

DATE/TIME OF TEST: 11/24/2017 01:53:26

Diagnostics Check	OK	01:50	
Air Blank	0.000	01:51	
Control Test	0.079	01:51	
Air Blank	0.000	01:51	
Subject Sample	# 1	01:52	Subject Test Refused
Breath Volume	0.000	01:52	
Air Blank	0.000	01:52	
Control Test	0.078	01:52	
Air Blank	0.000	01:53	
Diagnostics Check	OK	01:53	

Data Download Date/Time: 12/11/2017 15:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: STANLEY,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:58

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 11/24/2017 04:23:12

Diagnostics Check	OK	04:20	
Air Blank	0.000	04:20	
Control Test	0.079	04:21	
Air Blank	0.000	04:21	
Subject Sample	# 1	04:21	Subject Test Refused
Breath Volume	0.000	04:21	
Air Blank	0.000	04:22	
Control Test	0.078	04:22	
Air Blank	0.000	04:23	
Diagnostics Check	OK	04:23	

Data Download Date/Time: 12/11/2017 15:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAKER, TYLER, R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FISETTE, J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:35

SUBJECT'S LAST NAME: BELTTARI

DATE/TIME OF TEST: 11/24/2017 22:15:43

Diagnostics Check	OK	22:09
Air Blank	0.000	22:09
Control Test	0.079	22:10
Air Blank	0.000	22:10
Subject Sample # 1	0.000	22:11
Breath Volume	3.566	22:11
Air Blank	0.000	22:11
Air Blank	0.000	22:13
Subject Sample # 2	0.000	22:14
Breath Volume	2.539	22:14
Air Blank	0.000	22:14
Control Test	0.078	22:15
Air Blank	0.000	22:15
Diagnostics Check	OK	22:15

Data Download Date/Time: 12/11/2017 15:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON, BRADEN, J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BYRNE, J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: IRBY

DATE/TIME OF TEST: 11/30/2017 01:49:49

Diagnostics Check	OK	01:43
Air Blank	0.000	01:43
Control Test	0.079	01:43
Air Blank	0.000	01:44
Subject Sample # 1	0.202	01:45
Breath Volume	4.457	01:45
Air Blank	0.000	01:45
Air Blank	0.000	01:47
Subject Sample # 2	0.206	01:48
Breath Volume	4.468	01:48
Air Blank	0.000	01:48
Control Test	0.078	01:49
Air Blank	0.000	01:49
Diagnostics Check	OK	01:49

Data Download Date/Time: 12/11/2017 15:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON, BRADEN, J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: EANES, J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:43

SUBJECT'S LAST NAME: CLARK

DATE/TIME OF TEST: 11/30/2017 03:11:17

Diagnostics Check	OK	03:04
Air Blank	0.000	03:04
Control Test	0.079	03:04
Air Blank	0.000	03:05
Subject Sample # 1	0.207	03:06
Breath Volume	1.332	03:06
Air Blank	0.000	03:06
Air Blank	0.000	03:08
Subject Sample # 2	0.210	03:09
Breath Volume	1.636	03:09
Air Blank	0.000	03:10
Control Test	0.077	03:10
Air Blank	0.000	03:11
Diagnostics Check	OK	03:11

Data Download Date/Time: 12/11/2017 15:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLARIDA,K,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CHAPMAN,S
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 11:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:07

SUBJECT'S LAST NAME: VINING

DATE/TIME OF TEST: 11/30/2017 13:35:31

Diagnostics Check	OK	13:28
Air Blank	0.000	13:29
Control Test	0.078	13:29
Air Blank	0.000	13:29
Subject Sample # 1	0.037	13:30
Breath Volume	1.285	13:30
Air Blank	0.000	13:31
Air Blank	0.000	13:32
Subject Sample # 2	0.036	13:33
Breath Volume	1.585	13:33
Air Blank	0.000	13:34
Control Test	0.079	13:34
Air Blank	0.000	13:35
Diagnostics Check	OK	13:35

Data Download Date/Time: 12/11/2017 15:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSEPILER,JONATHAN,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: GARCIA,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:16

SUBJECT'S LAST NAME: BENNETT

DATE/TIME OF TEST: 11/30/2017 23:42:59

Diagnostics Check	OK	23:36
Air Blank	0.000	23:37
Control Test	0.078	23:37
Air Blank	0.000	23:37
Subject Sample # 1	0.209	23:38
Breath Volume	3.183	23:38
Air Blank	0.000	23:39
Air Blank	0.000	23:40
Subject Sample # 2	0.196	23:41
Breath Volume	3.656	23:41
Air Blank	0.000	23:42
Control Test	0.078	23:42
Air Blank	0.000	23:42
Diagnostics Check	OK	23:42

Data Download Date/Time: 12/11/2017 15:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001043

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROE,ROBERT,D
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: ROE,ROBERT
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 11/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: COURT ORD
ARREST TIME: 10:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:30

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 11/30/2017 11:20:28

Diagnostics Check	OK	11:14
Air Blank	0.000	11:14
Control Test	0.077	11:15
Air Blank	0.000	11:15
Subject Sample # 1	0.000	11:16
Breath Volume	2.851	11:16
Air Blank	0.000	11:16
Air Blank	0.000	11:18
Subject Sample # 2	0.000	11:19
Breath Volume	2.125	11:19
Air Blank	0.000	11:19
Control Test	0.077	11:19
Air Blank	0.000	11:20
Diagnostics Check	OK	11:20

Data Download Date/Time: 12/30/2017 09:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESCOBAR,ALEJANDRO,
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf
ARRESTING OFFICER: ESCOBAR,ALEJANDRO
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf
LAST AGENCY INSPECTION: 10/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 21:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: SILVA

DATE/TIME OF TEST: 11/03/2017 22:36:59

Diagnostics Check	OK	22:30
Air Blank	0.000	22:31
Control Test	0.079	22:31
Air Blank	0.000	22:31
Subject Sample # 1	0.269	22:32
Breath Volume	1.949	22:32
Air Blank	0.000	22:33
Air Blank	0.000	22:35
Subject Sample # 2	0.265	22:35
Breath Volume	2.257	22:35
Air Blank	0.000	22:36
Control Test	0.080	22:36
Air Blank	0.000	22:36
Diagnostics Check	OK	22:36

Data Download Date/Time: 11/28/2017 11:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESCOBAR,ALEJANDRO,
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf
ARRESTING OFFICER: ESCOBAR,ALEJANDRO
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf
LAST AGENCY INSPECTION: 10/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 02:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:41

SUBJECT'S LAST NAME: NELSON

DATE/TIME OF TEST: 11/10/2017 03:12:19

Diagnostics Check	OK	03:09	
Air Blank	0.000	03:09	
Control Test	0.080	03:10	
Air Blank	0.000	03:10	
Subject Sample	# 1	03:10	Subject Test Refused
Breath Volume	0.000	03:10	
Air Blank	0.000	03:11	
Control Test	0.080	03:11	
Air Blank	0.000	03:12	
Diagnostics Check	OK	03:12	

Data Download Date/Time: 11/28/2017 11:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESCOBAR,ALEJANDRO,
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf
ARRESTING OFFICER: WETHERINGTON,STEPHEN
ARRESTING OFFICER AGENCY: Coral Springs PD
LAST AGENCY INSPECTION: 10/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 02:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:33

SUBJECT'S LAST NAME: FERDINAND

DATE/TIME OF TEST: 11/11/2017 04:10:21

Diagnostics Check	OK	04:03
Air Blank	0.000	04:04
Control Test	0.079	04:04
Air Blank	0.000	04:04
Subject Sample # 1	0.166	04:05
Breath Volume	2.726	04:05
Air Blank	0.000	04:06
Air Blank	0.000	04:08
Subject Sample # 2	0.166	04:08
Breath Volume	1.886	04:08
Air Blank	0.000	04:09
Control Test	0.079	04:09
Air Blank	0.000	04:10
Diagnostics Check	OK	04:10

Data Download Date/Time: 11/28/2017 11:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LARA,JENNIFER,A
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf
ARRESTING OFFICER: DAVIS,DONALD
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf
LAST AGENCY INSPECTION: 10/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 10:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 11:00

SUBJECT'S LAST NAME: ROSERO

DATE/TIME OF TEST: 11/11/2017 11:34:09

Diagnostics Check	OK	11:29	
Air Blank	0.000	11:29	
Control Test	0.080	11:29	
Air Blank	0.000	11:30	
Subject Sample	# 1	11:32	Subject Test Refused
Breath Volume	0.000	11:32	
Air Blank	0.000	11:33	
Control Test	0.081	11:33	
Air Blank	0.000	11:34	
Diagnostics Check	OK	11:34	

Data Download Date/Time: 11/28/2017 11:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VASSILAKIS,GEORGE,E
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf
ARRESTING OFFICER: TANFULLA,DANTE
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf
LAST AGENCY INSPECTION: 10/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 20:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: SWAN

DATE/TIME OF TEST: 11/12/2017 23:06:15

Diagnostics Check	OK	23:03	
Air Blank	0.000	23:03	
Control Test	0.079	23:04	
Air Blank	0.000	23:04	
Subject Sample	# 1	23:04	Subject Test Refused
Breath Volume	0.000	23:04	
Air Blank	0.000	23:05	
Control Test	0.080	23:05	
Air Blank	0.000	23:06	
Diagnostics Check	OK	23:06	

Data Download Date/Time: 11/28/2017 11:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESCOBAR,ALEJANDRO,
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf
ARRESTING OFFICER: ESCOBAR,ALEJANDRO
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf
LAST AGENCY INSPECTION: 10/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 23:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 11/17/2017 23:54:55

Diagnostics Check	OK	23:51	
Air Blank	0.000	23:52	
Control Test	0.079	23:52	
Air Blank	0.000	23:53	
Subject Sample	# 1	23:53	Subject Test Refused
Breath Volume	0.000	23:53	
Air Blank	0.000	23:54	
Control Test	0.079	23:54	
Air Blank	0.000	23:54	
Diagnostics Check	OK	23:54	

Data Download Date/Time: 11/28/2017 11:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESCOBAR,ALEJANDRO,
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf
ARRESTING OFFICER: ESCOBAR,ALEJANDRO
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf
LAST AGENCY INSPECTION: 10/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 21:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:01

SUBJECT'S LAST NAME: KINCHEN

DATE/TIME OF TEST: 11/21/2017 22:36:43

Diagnostics Check	OK	22:30
Air Blank	0.000	22:30
Control Test	0.079	22:31
Air Blank	0.000	22:31
Subject Sample # 1	0.135	22:32
Breath Volume	2.207	22:32
Air Blank	0.000	22:32
Air Blank	0.000	22:34
Subject Sample # 2	0.134	22:35
Breath Volume	1.902	22:35
Air Blank	0.000	22:35
Control Test	0.080	22:36
Air Blank	0.000	22:36
Diagnostics Check	OK	22:36

Data Download Date/Time: 11/28/2017 11:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LARA,JENNIFER,A
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf
ARRESTING OFFICER: FLORES,CHRISTAN
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf
LAST AGENCY INSPECTION: 10/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 16:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:02

SUBJECT'S LAST NAME: BLAZER

DATE/TIME OF TEST: 11/26/2017 17:37:18

Diagnostics Check	OK	17:30
Air Blank	0.000	17:31
Control Test	0.079	17:31
Air Blank	0.000	17:32
Subject Sample # 1	0.187	17:33
Breath Volume	1.441	17:33
Air Blank	0.000	17:33
Air Blank	0.000	17:35
Subject Sample # 2	0.190	17:35
Breath Volume	2.136	17:35
Air Blank	0.000	17:36
Control Test	0.077	17:36
Air Blank	0.000	17:37
Diagnostics Check	OK	17:37

Data Download Date/Time: 11/28/2017 11:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VASSILAKIS,GEORGE,E
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf
ARRESTING OFFICER: VASSILAKIS,GEORGE
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf
LAST AGENCY INSPECTION: 11/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 22:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:05

SUBJECT'S LAST NAME: ROBINSON

DATE/TIME OF TEST: 11/29/2017 23:39:55

Diagnostics Check	OK	23:37	
Air Blank	0.000	23:37	
Control Test	0.079	23:37	
Air Blank	0.000	23:38	
Subject Sample	# 1	23:38	Subject Test Refused
Breath Volume	0.000	23:38	
Air Blank	0.000	23:39	
Control Test	0.080	23:39	
Air Blank	0.000	23:39	
Diagnostics Check	OK	23:39	

Data Download Date/Time: 12/27/2017 15:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001050

INSTRUMENT REGISTERED TO: CORAL SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROADHURST,RYAN,E
BREATH TEST OPERATOR AGENCY: Coral Springs PD
ARRESTING OFFICER: POPOCK,BRETT
ARRESTING OFFICER AGENCY: Coral Springs PD
LAST AGENCY INSPECTION: 10/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: BORGES

DATE/TIME OF TEST: 11/04/2017 00:34:01

Diagnostics Check	OK	00:31	
Air Blank	0.000	00:31	
Control Test	0.080	00:31	
Air Blank	0.000	00:32	
Subject Sample	# 1	00:32	Subject Test Refused
Breath Volume	0.000	00:32	
Air Blank	0.000	00:33	
Control Test	0.080	00:33	
Air Blank	0.000	00:33	
Diagnostics Check	OK	00:34	

Data Download Date/Time: 11/08/2017 08:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOWE, WILLIAM, E
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: ADMIN, ADMIN
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 11/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: ADMIN
ARREST TIME: 13:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:40

SUBJECT'S LAST NAME: LOWE

DATE/TIME OF TEST: 11/02/2017 14:24:27

Diagnostics Check	OK	14:13	
Air Blank	0.000	14:13	
Control Test	0.081	14:14	
Air Blank	0.000	14:14	
Subject Sample	# 1 0.171	14:15	
Breath Volume	1.582	14:15	
Air Blank	0.000	14:16	
Air Blank	0.000	14:18	
Subject Sample	# 2 0.195	14:18	
Breath Volume	1.738	14:18	
Air Blank	0.000	14:19	
Air Blank	0.000	14:21	
Subject Sample	# 3	14:23	Subject Test Refused
Breath Volume	0.000	14:23	
Air Blank	0.000	14:23	
Control Test	0.080	14:23	
Air Blank	0.000	14:24	
Diagnostics Check	OK	14:24	

Data Download Date/Time: 12/14/2017 07:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HODGES,P,E
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: HODGES,P
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 11/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: ADMIN
ARREST TIME: 14:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:31

SUBJECT'S LAST NAME: OQUINN

DATE/TIME OF TEST: 11/02/2017 14:59:40

Diagnostics Check	OK	14:51	
Air Blank	0.000	14:52	
Control Test	0.080	14:52	
Air Blank	0.000	14:53	
Subject Sample # 1	0.123	14:53	
Breath Volume	1.503	14:53	
Air Blank	0.000	14:54	
Air Blank	0.000	14:56	
Subject Sample # 2		14:58	Subject Test Refused
Breath Volume	0.156	14:58	
Air Blank	0.000	14:58	
Control Test	0.081	14:59	
Air Blank	0.000	14:59	
Diagnostics Check	OK	14:59	

Data Download Date/Time: 12/14/2017 07:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HODGES,P,E
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: HODGES,P
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 11/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: ADMIN
ARREST TIME: 14:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:30

SUBJECT'S LAST NAME: HORTON

DATE/TIME OF TEST: 11/02/2017 15:11:49

Diagnostics Check	OK	15:05
Air Blank	0.000	15:05
Control Test	0.080	15:05
Air Blank	0.000	15:06
Subject Sample # 1	0.064	15:07
Breath Volume	1.578	15:07
Air Blank	0.000	15:07
Air Blank	0.000	15:09
Subject Sample # 2	0.064	15:10
Breath Volume	1.203	15:10
Air Blank	0.000	15:10
Control Test	0.081	15:11
Air Blank	0.000	15:11
Diagnostics Check	OK	15:11

Data Download Date/Time: 12/14/2017 07:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HODGES,P,E
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: HODGES,P
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 11/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: ADMIN
ARREST TIME: 14:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:30

SUBJECT'S LAST NAME: GOTTULA

DATE/TIME OF TEST: 11/02/2017 15:20:25

Diagnostics Check	OK	15:13
Air Blank	0.000	15:13
Control Test	0.081	15:14
Air Blank	0.000	15:14
Subject Sample # 1	0.093	15:15
Breath Volume	1.433	15:15
Air Blank	0.000	15:16
Air Blank	0.000	15:18
Subject Sample # 2	0.093	15:18
Breath Volume	1.265	15:18
Air Blank	0.000	15:19
Control Test	0.080	15:19
Air Blank	0.000	15:20
Diagnostics Check	OK	15:20

Data Download Date/Time: 12/14/2017 07:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPBELL,D,R
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: THOLL,K
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 11/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: RICHTER

DATE/TIME OF TEST: 11/03/2017 23:37:20

Diagnostics Check	OK	23:30
Air Blank	0.000	23:31
Control Test	0.081	23:31
Air Blank	0.000	23:32
Subject Sample # 1	0.217	23:32
Breath Volume	1.140	23:32
Air Blank	0.000	23:33
Air Blank	0.000	23:35
Subject Sample # 2	0.224	23:35
Breath Volume	1.363	23:35
Air Blank	0.000	23:36
Control Test	0.079	23:36
Air Blank	0.000	23:37
Diagnostics Check	OK	23:37

Data Download Date/Time: 12/14/2017 07:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARDONA,KALE,C
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: ZITTROWER,T
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 11/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:09

SUBJECT'S LAST NAME: NETTLES

DATE/TIME OF TEST: 11/05/2017 01:44:30

Diagnostics Check	OK	01:41	
Air Blank	0.000	01:41	
Control Test	0.080	01:42	
Air Blank	0.000	01:42	
Subject Sample	# 1	01:43	Subject Test Refused
Breath Volume	0.000	01:43	
Air Blank	0.000	01:43	
Control Test	0.080	01:43	
Air Blank	0.000	01:44	
Diagnostics Check	OK	01:44	

Data Download Date/Time: 12/14/2017 07:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARDON,K,C
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: RENYOLDS,M
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 11/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: SKINNER

DATE/TIME OF TEST: 11/09/2017 23:32:22

Diagnostics Check	OK	23:20	
Air Blank	0.000	23:21	
Control Test	0.080	23:21	
Air Blank	0.000	23:22	
Subject Sample	# 1 0.184	23:25	Volume Not Met
Breath Volume	0.019	23:25	
Air Blank	0.000	23:26	
Air Blank	0.000	23:27	
Subject Sample	# 2 0.188	23:30	Volume Not Met
Breath Volume	0.113	23:30	
Air Blank	0.000	23:31	
Control Test	0.079	23:31	
Air Blank	0.000	23:32	
Diagnostics Check	OK	23:32	

Data Download Date/Time: 12/14/2017 07:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THORNTON,N,E
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: ENNIS,J
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 11/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 18:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:53

SUBJECT'S LAST NAME: BUNCH

DATE/TIME OF TEST: 11/15/2017 19:21:15

Diagnostics Check	OK	19:14
Air Blank	0.000	19:14
Control Test	0.080	19:14
Air Blank	0.000	19:15
Subject Sample # 1	0.204	19:16
Breath Volume	1.929	19:16
Air Blank	0.000	19:16
Air Blank	0.000	19:18
Subject Sample # 2	0.201	19:19
Breath Volume	1.437	19:19
Air Blank	0.000	19:20
Control Test	0.081	19:20
Air Blank	0.000	19:21
Diagnostics Check	OK	19:21

Data Download Date/Time: 12/14/2017 07:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARLEY,JOSEPH,E
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: FARLEY,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 11/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: CORBITT

DATE/TIME OF TEST: 11/19/2017 01:17:14

Diagnostics Check	OK	01:10
Air Blank	0.000	01:11
Control Test	0.081	01:11
Air Blank	0.000	01:12
Subject Sample # 1	0.158	01:12
Breath Volume	2.460	01:12
Air Blank	0.000	01:13
Air Blank	0.000	01:15
Subject Sample # 2	0.160	01:15
Breath Volume	2.472	01:15
Air Blank	0.000	01:16
Control Test	0.081	01:16
Air Blank	0.000	01:17
Diagnostics Check	OK	01:17

Data Download Date/Time: 12/14/2017 07:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRES,J,L
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: HUNTER,M
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 11/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:02

SUBJECT'S LAST NAME: OURS

DATE/TIME OF TEST: 11/21/2017 22:25:20

Diagnostics Check	OK	22:22	
Air Blank	0.000	22:22	
Control Test	0.080	22:23	
Air Blank	0.000	22:23	
Subject Sample	# 1	22:23	Subject Test Refused
Breath Volume	0.000	22:23	
Air Blank	0.000	22:24	
Control Test	0.080	22:24	
Air Blank	0.000	22:25	
Diagnostics Check	OK	22:25	

Data Download Date/Time: 12/14/2017 07:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THORNTON,N,E
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: CONE,A
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 11/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:17

SUBJECT'S LAST NAME: OQUENDO

DATE/TIME OF TEST: 11/26/2017 01:45:58

Diagnostics Check	OK	01:43	
Air Blank	0.000	01:43	
Control Test	0.080	01:43	
Air Blank	0.000	01:44	
Subject Sample	# 1	01:44	Subject Test Refused
Breath Volume	0.000	01:44	
Air Blank	0.000	01:45	
Control Test	0.081	01:45	
Air Blank	0.000	01:45	
Diagnostics Check	OK	01:45	

Data Download Date/Time: 12/14/2017 07:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARLOW,ROYCE,B
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: SALVADOR,ELDER
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 10/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:06

SUBJECT'S LAST NAME: OTERO

DATE/TIME OF TEST: 11/15/2017 03:30:28

Diagnostics Check	OK	03:27	
Air Blank	0.000	03:28	
Control Test	0.080	03:28	
Air Blank	0.000	03:28	
Subject Sample	# 1	03:29	Subject Test Refused
Breath Volume	0.000	03:29	
Air Blank	0.000	03:29	
Control Test	0.080	03:29	
Air Blank	0.000	03:30	
Diagnostics Check	OK	03:30	

Data Download Date/Time: 11/28/2017 08:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SVENDSEN,C,E
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: SVENDSEN,C
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 10/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: OTHER
ARREST TIME: 11:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:45

SUBJECT'S LAST NAME: GRANT

DATE/TIME OF TEST: 11/15/2017 12:24:22

Diagnostics Check	OK	12:18
Air Blank	0.000	12:18
Control Test	0.079	12:18
Air Blank	0.000	12:19
Subject Sample # 1	0.000	12:20
Breath Volume	1.914	12:20
Air Blank	0.000	12:20
Air Blank	0.000	12:22
Subject Sample # 2	0.000	12:23
Breath Volume	2.269	12:23
Air Blank	0.000	12:23
Control Test	0.080	12:23
Air Blank	0.000	12:24
Diagnostics Check	OK	12:24

Data Download Date/Time: 11/28/2017 08:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PECK,JOHN,R
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: PECK,JOHN
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 10/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: OTHER
ARREST TIME: 13:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:00

SUBJECT'S LAST NAME: BOYD

DATE/TIME OF TEST: 11/16/2017 13:43:21

Diagnostics Check	OK	13:36
Air Blank	0.000	13:37
Control Test	0.079	13:37
Air Blank	0.000	13:38
Subject Sample # 1	0.000	13:38
Breath Volume	1.703	13:38
Air Blank	0.000	13:39
Air Blank	0.000	13:41
Subject Sample # 2	0.000	13:41
Breath Volume	1.781	13:41
Air Blank	0.000	13:42
Control Test	0.080	13:42
Air Blank	0.000	13:43
Diagnostics Check	OK	13:43

Data Download Date/Time: 11/28/2017 08:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERRANTE,TRACY,L
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: SALVADOR,ELDER
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 10/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:04

SUBJECT'S LAST NAME: MCKINNIE

DATE/TIME OF TEST: 11/19/2017 04:37:07

Diagnostics Check	OK	04:30
Air Blank	0.000	04:30
Control Test	0.081	04:30
Air Blank	0.000	04:31
Subject Sample # 1	0.117	04:32
Breath Volume	2.421	04:32
Air Blank	0.000	04:33
Air Blank	0.000	04:35
Subject Sample # 2	0.112	04:35
Breath Volume	2.222	04:35
Air Blank	0.000	04:36
Control Test	0.078	04:36
Air Blank	0.000	04:37
Diagnostics Check	OK	04:37

Data Download Date/Time: 11/28/2017 08:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ABELN,SHON,W
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: ABELN,SHON
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 10/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: COURT ORD
ARREST TIME: 13:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:00

SUBJECT'S LAST NAME: GRANT

DATE/TIME OF TEST: 11/20/2017 13:36:56

Diagnostics Check	OK	13:31
Air Blank	0.000	13:31
Control Test	0.079	13:31
Air Blank	0.000	13:32
Subject Sample # 1	0.000	13:32
Breath Volume	1.808	13:32
Air Blank	0.000	13:33
Air Blank	0.000	13:35
Subject Sample # 2	0.000	13:35
Breath Volume	1.894	13:35
Air Blank	0.000	13:36
Control Test	0.080	13:36
Air Blank	0.000	13:36
Diagnostics Check	OK	13:36

Data Download Date/Time: 11/28/2017 08:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GLASCO,GREGORY,S
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: GLASCO,GREGORY
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 10/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: COURT ORD
ARREST TIME: 13:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:58

SUBJECT'S LAST NAME: MOISTNER

DATE/TIME OF TEST: 11/22/2017 14:33:22

Diagnostics Check	OK	14:28	
Air Blank	0.000	14:28	
Control Test	0.080	14:29	
Air Blank	0.000	14:29	
Subject Sample # 1	0.000	14:29	
Breath Volume	2.167	14:29	
Air Blank	0.000	14:30	
Air Blank	0.000	14:32	
Subject Sample # 2	0.000	14:32	
Breath Volume	2.367	14:32	
Air Blank		14:32	RFI Detect
Air Blank	0.000	14:33	

Data Download Date/Time: 11/28/2017 08:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GLASCO,GREGORY,S
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: GLASCO,GREGORY
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 10/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: COURT ORD
ARREST TIME: 13:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:58

SUBJECT'S LAST NAME: MOISTNER

DATE/TIME OF TEST: 11/22/2017 14:41:34

Diagnostics Check	OK	14:35
Air Blank	0.000	14:36
Control Test	0.079	14:36
Air Blank	0.000	14:37
Subject Sample # 1	0.000	14:37
Breath Volume	1.804	14:37
Air Blank	0.000	14:37
Air Blank	0.000	14:39
Subject Sample # 2	0.000	14:40
Breath Volume	1.882	14:40
Air Blank	0.000	14:40
Control Test	0.080	14:40
Air Blank	0.000	14:41
Diagnostics Check	OK	14:41

Data Download Date/Time: 11/28/2017 08:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,NAKEEM,T
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: BROWN,NAKEEM
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 10/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: COURT ORD
ARREST TIME: 14:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:20

SUBJECT'S LAST NAME: MCTAGGART

DATE/TIME OF TEST: 11/22/2017 15:00:07

Diagnostics Check	OK	14:53
Air Blank	0.000	14:54
Control Test	0.080	14:54
Air Blank	0.000	14:55
Subject Sample # 1	0.000	14:55
Breath Volume	1.414	14:55
Air Blank	0.000	14:56
Air Blank	0.000	14:58
Subject Sample # 2	0.000	14:58
Breath Volume	1.539	14:58
Air Blank	0.000	14:59
Control Test	0.080	14:59
Air Blank	0.000	15:00
Diagnostics Check	OK	15:00

Data Download Date/Time: 11/28/2017 08:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMILTON,COURTNEY,L
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: SALVADOR,ELDER
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 10/21/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:53

SUBJECT'S LAST NAME: SAUL

DATE/TIME OF TEST: 11/23/2017 01:20:28

Diagnostics Check	OK	01:14
Air Blank	0.000	01:14
Control Test	0.079	01:14
Air Blank	0.000	01:15
Subject Sample # 1	0.166	01:15
Breath Volume	3.125	01:15
Air Blank	0.000	01:16
Air Blank	0.000	01:18
Subject Sample # 2	0.167	01:18
Breath Volume	4.187	01:18
Air Blank	0.000	01:19
Control Test	0.079	01:19
Air Blank	0.000	01:20
Diagnostics Check	OK	01:20

Data Download Date/Time: 11/28/2017 08:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERRANTE,TRACY,L
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: FERRANTE,TRACY
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 10/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 12:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:40

SUBJECT'S LAST NAME: GRANT

DATE/TIME OF TEST: 11/23/2017 13:07:25

Diagnostics Check	OK	13:01
Air Blank	0.000	13:02
Control Test	0.080	13:02
Air Blank	0.000	13:02
Subject Sample # 1	0.000	13:03
Breath Volume	1.882	13:03
Air Blank	0.000	13:03
Air Blank	0.000	13:05
Subject Sample # 2	0.000	13:06
Breath Volume	2.296	13:06
Air Blank	0.000	13:06
Control Test	0.079	13:06
Air Blank	0.000	13:07
Diagnostics Check	OK	13:07

Data Download Date/Time: 11/28/2017 08:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINISH,BARRY,A
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: SALVADOR,ELDER
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 10/21/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:49

SUBJECT'S LAST NAME: ALLEN

DATE/TIME OF TEST: 11/24/2017 05:21:00

Diagnostics Check	OK	05:18	
Air Blank	0.000	05:18	
Control Test	0.080	05:18	
Air Blank	0.000	05:19	
Subject Sample	# 1	05:19	Subject Test Refused
Breath Volume	0.000	05:19	
Air Blank	0.000	05:20	
Control Test	0.079	05:20	
Air Blank	0.000	05:20	
Diagnostics Check	OK	05:21	

Data Download Date/Time: 11/28/2017 08:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ABELN,SHON,W
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: ABELN,SHON
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 10/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: COURT ORD
ARREST TIME: 12:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 12:30

SUBJECT'S LAST NAME: DEBORAH

DATE/TIME OF TEST: 11/26/2017 13:00:07

Diagnostics Check	OK	12:54
Air Blank	0.000	12:54
Control Test	0.080	12:55
Air Blank	0.000	12:55
Subject Sample # 1	0.000	12:55
Breath Volume	1.828	12:55
Air Blank	0.000	12:56
Air Blank	0.000	12:58
Subject Sample # 2	0.000	12:58
Breath Volume	1.730	12:58
Air Blank	0.000	12:59
Control Test	0.080	12:59
Air Blank	0.000	13:00
Diagnostics Check	OK	13:00

Data Download Date/Time: 11/28/2017 08:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERRANTE,TRACY,L
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: FERRANTE,TRACY
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 10/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 12:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:45

SUBJECT'S LAST NAME: GRANT

DATE/TIME OF TEST: 11/27/2017 13:16:35

Diagnostics Check	OK	13:10
Air Blank	0.000	13:11
Control Test	0.080	13:11
Air Blank	0.000	13:11
Subject Sample # 1	0.000	13:12
Breath Volume	1.648	13:12
Air Blank	0.000	13:12
Air Blank	0.000	13:14
Subject Sample # 2	0.000	13:15
Breath Volume	2.152	13:15
Air Blank	0.000	13:15
Control Test	0.080	13:15
Air Blank	0.000	13:16
Diagnostics Check	OK	13:16

Data Download Date/Time: 11/28/2017 08:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNELL,C,J
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: SALVADOR,E
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 11/28/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 12:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:42

SUBJECT'S LAST NAME: DAGES

DATE/TIME OF TEST: 11/29/2017 02:09:28

Diagnostics Check	OK	02:02
Air Blank	0.000	02:03
Control Test	0.080	02:03
Air Blank	0.000	02:03
Subject Sample # 1	0.094	02:04
Breath Volume	2.289	02:04
Air Blank	0.000	02:05
Air Blank	0.000	02:07
Subject Sample # 2	0.089	02:07
Breath Volume	1.652	02:07
Air Blank	0.000	02:08
Control Test	0.081	02:08
Air Blank	0.000	02:09
Diagnostics Check	OK	02:09

Data Download Date/Time: 12/30/2017 08:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WORTH,KEVIN,J
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: DENNIS,A
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 11/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: AGUILLON

DATE/TIME OF TEST: 11/30/2017 00:06:23

Diagnostics Check	OK	00:00
Air Blank	0.000	00:00
Control Test	0.080	00:01
Air Blank	0.000	00:01
Subject Sample # 1	0.220	00:02
Breath Volume	2.406	00:02
Air Blank	0.000	00:02
Air Blank	0.000	00:04
Subject Sample # 2	0.229	00:04
Breath Volume	2.335	00:04
Air Blank	0.000	00:05
Control Test	0.079	00:05
Air Blank	0.000	00:06
Diagnostics Check	OK	00:06

Data Download Date/Time: 12/30/2017 08:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GETSEE,JOY,S
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: WOFFORD,CLINT
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 10/09/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080AI
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 17:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:25

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 11/01/2017 18:55:56

Diagnostics Check	OK	18:49
Air Blank	0.000	18:50
Control Test	0.077	18:50
Air Blank	0.000	18:51
Subject Sample # 1	0.168	18:51
Breath Volume	1.792	18:51
Air Blank	0.000	18:52
Air Blank	0.000	18:53
Subject Sample # 2	0.160	18:54
Breath Volume	1.406	18:54
Air Blank	0.000	18:55
Control Test	0.078	18:55
Air Blank	0.000	18:55
Diagnostics Check	OK	18:55

Data Download Date/Time: 11/12/2017 18:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUCAS,NATHANIEL,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: LUCAS,NATHANIEL
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 10/09/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080AI
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: MCKAY

DATE/TIME OF TEST: 11/01/2017 23:46:18

Diagnostics Check	OK	23:40
Air Blank	0.000	23:40
Control Test	0.080	23:41
Air Blank	0.000	23:41
Subject Sample # 1	0.000	23:42
Breath Volume	3.488	23:42
Air Blank	0.000	23:42
Air Blank	0.000	23:44
Subject Sample # 2	0.000	23:44
Breath Volume	3.316	23:44
Air Blank	0.000	23:45
Control Test	0.080	23:45
Air Blank	0.000	23:46
Diagnostics Check	OK	23:46

Data Download Date/Time: 11/12/2017 18:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HEWETT,MICHAEL,W
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: HEWETT,MICHAEL
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 10/09/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080AI
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 15:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:40

SUBJECT'S LAST NAME: BYRNE

DATE/TIME OF TEST: 11/02/2017 16:11:19

Diagnostics Check	OK	16:05
Air Blank	0.000	16:05
Control Test	0.080	16:05
Air Blank	0.000	16:06
Subject Sample # 1	0.157	16:06
Breath Volume	1.820	16:06
Air Blank	0.000	16:07
Air Blank	0.000	16:09
Subject Sample # 2	0.161	16:09
Breath Volume	2.363	16:09
Air Blank	0.000	16:10
Control Test	0.081	16:10
Air Blank	0.000	16:11
Diagnostics Check	OK	16:11

Data Download Date/Time: 11/12/2017 18:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GETSEE,JOY,S
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: WEBSTER,BRAD
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 10/09/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080AI
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:54

SUBJECT'S LAST NAME: PATEL

DATE/TIME OF TEST: 11/03/2017 01:21:09

Diagnostics Check	OK	01:17	
Air Blank	0.000	01:18	
Control Test	0.080	01:18	
Air Blank	0.000	01:19	
Subject Sample	# 1	01:19	Subject Test Refused
Breath Volume	0.000	01:19	
Air Blank	0.000	01:20	
Control Test	0.081	01:20	
Air Blank	0.000	01:21	
Diagnostics Check	OK	01:21	

Data Download Date/Time: 11/12/2017 18:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GETSEE,JOY,S
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: WOFFORD,CLINT
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 10/09/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080AI
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: BENNETT

DATE/TIME OF TEST: 11/03/2017 02:21:19

Diagnostics Check	OK	02:18	
Air Blank	0.000	02:18	
Control Test	0.081	02:19	
Air Blank	0.000	02:19	
Subject Sample	# 1	02:19	Subject Test Refused
Breath Volume	0.000	02:19	
Air Blank	0.000	02:20	
Control Test	0.080	02:20	
Air Blank	0.000	02:21	
Diagnostics Check	OK	02:21	

Data Download Date/Time: 11/12/2017 18:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OGG,DAVID,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: OGG,DAVID
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 10/09/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080AI
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 14:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:48

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 11/03/2017 15:14:54

Diagnostics Check	OK	15:08
Air Blank	0.000	15:08
Control Test	0.081	15:09
Air Blank	0.000	15:09
Subject Sample # 1	0.000	15:10
Breath Volume	1.703	15:10
Air Blank	0.000	15:10
Air Blank	0.000	15:12
Subject Sample # 2	0.000	15:13
Breath Volume	1.789	15:13
Air Blank	0.000	15:14
Control Test	0.081	15:14
Air Blank	0.000	15:14
Diagnostics Check	OK	15:14

Data Download Date/Time: 11/12/2017 18:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARREN,BILLY,J
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: WARREN,BILLY
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 10/09/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080AI
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:10

SUBJECT'S LAST NAME: OLIVER

DATE/TIME OF TEST: 11/04/2017 09:00:04

Diagnostics Check	OK	08:57	
Air Blank	0.000	08:57	
Control Test	0.080	08:58	
Air Blank	0.000	08:58	
Subject Sample	# 1	08:58	Subject Test Refused
Breath Volume	0.000	08:58	
Air Blank	0.000	08:59	
Control Test	0.080	08:59	
Air Blank	0.000	08:59	
Diagnostics Check	OK	09:00	

Data Download Date/Time: 11/12/2017 18:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SINCLAIR,CAMERON,W
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: SINCLAIR,CAMERON
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 10/09/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080AI
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:58

SUBJECT'S LAST NAME: RABER

DATE/TIME OF TEST: 11/05/2017 00:31:50

Diagnostics Check	OK	00:20	
Air Blank	0.000	00:20	
Control Test	0.080	00:21	
Air Blank	0.000	00:21	
Subject Sample	# 1	00:24	No Sample Provided
Breath Volume	0.000	00:24	
Air Blank	0.000	00:25	
Air Blank	0.000	00:27	
Subject Sample	# 2	00:30	No Sample Provided
Breath Volume	0.000	00:30	
Air Blank	0.000	00:30	
Control Test	0.080	00:31	
Air Blank	0.000	00:31	
Diagnostics Check	OK	00:31	

Data Download Date/Time: 11/12/2017 18:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARREN,BILLY,J
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: WARREN,BILLY
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 10/09/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080AI
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:06

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 11/05/2017 01:42:31

Diagnostics Check	OK	01:32	
Air Blank	0.000	01:33	
Control Test	0.080	01:33	
Air Blank	0.000	01:34	
Subject Sample	# 1	01:37	No Sample Provided
Breath Volume	0.000	01:37	
Air Blank	0.000	01:37	
Air Blank	0.000	01:39	
Subject Sample	# 2	01:41	Subject Test Refused
Breath Volume	0.000	01:41	
Air Blank	0.000	01:41	
Control Test	0.080	01:41	
Air Blank	0.000	01:42	
Diagnostics Check	OK	01:42	

Data Download Date/Time: 11/12/2017 18:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAWES,ERIC,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: LAWES,ERIC
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 10/09/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080AI
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:39

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 11/06/2017 03:13:19

Diagnostics Check	OK	03:06
Air Blank	0.000	03:07
Control Test	0.081	03:07
Air Blank	0.000	03:08
Subject Sample # 1	0.195	03:08
Breath Volume	2.246	03:08
Air Blank	0.000	03:09
Air Blank	0.000	03:11
Subject Sample # 2	0.195	03:11
Breath Volume	2.644	03:11
Air Blank	0.000	03:12
Control Test	0.080	03:12
Air Blank	0.000	03:13
Diagnostics Check	OK	03:13

Data Download Date/Time: 11/12/2017 18:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SYKES,STEVEN,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: SYKES,STEVEN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 10/09/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080AI
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 18:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:38

SUBJECT'S LAST NAME: WEEMS JR

DATE/TIME OF TEST: 11/06/2017 19:12:38

Diagnostics Check	OK	19:00	
Air Blank	0.000	19:00	
Control Test	0.080	19:00	
Air Blank	0.000	19:01	
Subject Sample	# 1 0.060	19:04	Volume Not Met
Breath Volume	0.406	19:04	
Air Blank	0.000	19:05	
Air Blank	0.000	19:06	
Subject Sample	# 2 0.054	19:07	
Breath Volume	1.597	19:07	
Air Blank	0.000	19:08	
Air Blank	0.000	19:09	
Subject Sample	# 3 0.054	19:11	
Breath Volume	1.277	19:11	
Air Blank	0.000	19:11	
Control Test	0.080	19:12	
Air Blank	0.000	19:12	
Diagnostics Check	OK	19:12	

Data Download Date/Time: 11/12/2017 18:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COKE,CLEVELAND,A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: WOFFORD,CLINT
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 10/09/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080AI
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 18:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:30

SUBJECT'S LAST NAME: MUNOZ

DATE/TIME OF TEST: 11/07/2017 19:10:24

Diagnostics Check	OK	19:04
Air Blank	0.000	19:04
Control Test	0.081	19:04
Air Blank	0.000	19:05
Subject Sample # 1	0.089	19:05
Breath Volume	1.726	19:05
Air Blank	0.000	19:06
Air Blank	0.000	19:08
Subject Sample # 2	0.103	19:08
Breath Volume	1.984	19:08
Air Blank	0.000	19:09
Control Test	0.080	19:09
Air Blank	0.000	19:10
Diagnostics Check	OK	19:10

Data Download Date/Time: 11/12/2017 18:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAWES,ERIC,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: WOFFORD,CLINT
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 10/09/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080AI
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: GUNTHER

DATE/TIME OF TEST: 11/11/2017 01:12:48

Diagnostics Check	OK	01:06
Air Blank	0.000	01:06
Control Test	0.080	01:07
Air Blank	0.000	01:07
Subject Sample # 1	0.126	01:08
Breath Volume	1.468	01:08
Air Blank	0.000	01:08
Air Blank	0.000	01:10
Subject Sample # 2	0.113	01:11
Breath Volume	1.281	01:11
Air Blank	0.000	01:11
Control Test	0.080	01:12
Air Blank	0.000	01:12
Diagnostics Check	OK	01:12

Data Download Date/Time: 11/12/2017 18:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUCAS,NATHANIEL,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: TINDLE,ERIC
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 10/09/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080AI
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:24

SUBJECT'S LAST NAME: MELENDEZ

DATE/TIME OF TEST: 11/11/2017 02:46:41

Diagnostics Check	OK	02:45	
Air Blank	0.000	02:45	
Control Test	0.081	02:45	
Air Blank		02:46	RFI Detect
Air Blank	0.000	02:46	

Data Download Date/Time: 11/12/2017 18:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUCAS,NATHANIEL,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: TINDLE,ERIC
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 10/09/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080AI
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:24

SUBJECT'S LAST NAME: MELENDEZ

DATE/TIME OF TEST: 11/11/2017 02:55:59

Diagnostics Check	OK	02:48
Air Blank	0.000	02:49
Control Test	0.081	02:49
Air Blank	0.000	02:50
Subject Sample # 1	0.000	02:51
Breath Volume	1.367	02:51
Air Blank	0.000	02:52
Air Blank	0.000	02:54
Subject Sample # 2	0.000	02:54
Breath Volume	1.773	02:54
Air Blank	0.000	02:55
Control Test	0.080	02:55
Air Blank	0.000	02:55
Diagnostics Check	OK	02:55

Data Download Date/Time: 11/12/2017 18:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAWES,ERIC,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: WEBSTER,BRAD
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 10/09/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080AI
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:57

SUBJECT'S LAST NAME: STEHENS

DATE/TIME OF TEST: 11/11/2017 03:33:20

Diagnostics Check	OK	03:26
Air Blank	0.000	03:26
Control Test	0.080	03:26
Air Blank	0.000	03:27
Subject Sample # 1	0.021	03:28
Breath Volume	2.425	03:28
Air Blank	0.000	03:29
Air Blank	0.000	03:31
Subject Sample # 2	0.020	03:31
Breath Volume	2.750	03:31
Air Blank	0.000	03:32
Control Test	0.080	03:32
Air Blank	0.000	03:33
Diagnostics Check	OK	03:33

Data Download Date/Time: 11/12/2017 18:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANDERSON, DAVID, A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: BLACK, DAVID
ARRESTING OFFICER AGENCY: Lake Wales PD
LAST AGENCY INSPECTION: 10/09/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080AI
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 13:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:30

SUBJECT'S LAST NAME: CASTILLO

DATE/TIME OF TEST: 11/11/2017 15:43:15

Diagnostics Check	OK	15:40	
Air Blank	0.000	15:40	
Control Test	0.080	15:40	
Air Blank	0.000	15:41	
Subject Sample	# 1	15:41	Subject Test Refused
Breath Volume	0.000	15:41	
Air Blank	0.000	15:42	
Control Test	0.080	15:42	
Air Blank	0.000	15:43	
Diagnostics Check	OK	15:43	

Data Download Date/Time: 11/12/2017 18:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUCAS,NATHANIEL,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: LUCAS,NATHANIEL
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 10/09/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080AI
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: GALLEGOS-BIAS

DATE/TIME OF TEST: 11/11/2017 23:59:36

Diagnostics Check	OK	23:48	
Air Blank	0.000	23:48	
Control Test	0.080	23:49	
Air Blank	0.000	23:49	
Subject Sample	# 1	23:52	No Sample Provided
Breath Volume	0.000	23:52	
Air Blank	0.000	23:53	
Air Blank	0.000	23:55	
Subject Sample	# 2	23:58	No Sample Provided
Breath Volume	0.000	23:58	
Air Blank	0.000	23:58	
Control Test	0.080	23:59	
Air Blank	0.000	23:59	
Diagnostics Check	OK	23:59	

Data Download Date/Time: 11/12/2017 18:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUCAS,NATHANIEL,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: WOFFORD,CLINT
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 10/09/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080AI
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: BURDETTE

DATE/TIME OF TEST: 11/12/2017 02:52:08

Diagnostics Check	OK	02:40	
Air Blank	0.000	02:41	
Control Test	0.080	02:41	
Air Blank	0.000	02:42	
Subject Sample	# 1	02:45	No Sample Provided
Breath Volume	0.000	02:45	
Air Blank	0.000	02:45	
Air Blank	0.000	02:47	
Subject Sample	# 2	02:50	No Sample Provided
Breath Volume	0.000	02:50	
Air Blank	0.000	02:51	
Control Test	0.080	02:51	
Air Blank	0.000	02:52	
Diagnostics Check	OK	02:52	

Data Download Date/Time: 11/12/2017 18:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JEUDY,RINALDO,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: WOFFORD,CLINT
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080AI
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:50

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 11/12/2017 20:44:29

Diagnostics Check	OK	20:27	
Air Blank	0.000	20:28	
Control Test	0.081	20:28	
Air Blank	0.000	20:28	
Subject Sample	# 1 0.316	20:31	
Breath Volume	1.167	20:31	
Air Blank	0.000	20:32	
Air Blank	0.000	20:34	
Subject Sample	# 2 0.261	20:37	Volume Not Met
Breath Volume	0.082	20:37	
Air Blank	0.000	20:37	
Air Blank	0.000	20:39	
Subject Sample	# 3 0.236	20:42	Volume Not Met
Breath Volume	0.808	20:42	
Air Blank	0.000	20:43	
Control Test	0.078	20:43	
Air Blank	0.000	20:44	
Diagnostics Check	OK	20:44	

Data Download Date/Time: 12/23/2017 19:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUABA,JASON,T
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: GUABA,JASON
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080AI
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:09

SUBJECT'S LAST NAME: CORRE FERNANDEZ

DATE/TIME OF TEST: 11/12/2017 21:41:15

Diagnostics Check	OK	21:29	
Air Blank	0.000	21:29	
Control Test	0.079	21:30	
Air Blank	0.000	21:30	
Subject Sample	# 1	21:33	No Sample Provided
Breath Volume	0.000	21:33	
Air Blank	0.000	21:34	
Air Blank	0.000	21:36	
Subject Sample	# 2	0.149	21:39
Breath Volume	0.679	21:39	Volume Not Met
Air Blank	0.000	21:40	
Control Test	0.080	21:40	
Air Blank	0.000	21:41	
Diagnostics Check	OK	21:41	

Data Download Date/Time: 12/23/2017 19:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUCAS,NATHANIEL,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: LUCAS,NATHANIEL
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080AI
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: CRUZ

DATE/TIME OF TEST: 11/13/2017 03:40:49

Diagnostics Check	OK	03:34
Air Blank	0.000	03:35
Control Test	0.080	03:35
Air Blank	0.000	03:36
Subject Sample # 1	0.000	03:36
Breath Volume	1.464	03:36
Air Blank	0.000	03:37
Air Blank	0.000	03:39
Subject Sample # 2	0.000	03:39
Breath Volume	1.359	03:39
Air Blank	0.000	03:39
Control Test	0.081	03:40
Air Blank	0.000	03:40
Diagnostics Check	OK	03:40

Data Download Date/Time: 12/23/2017 19:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MELTZER,CHRIS,J
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: MELTZER,CHRIS
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080AI
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 11:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:35

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 11/13/2017 12:13:02

Diagnostics Check	OK	12:10	
Air Blank	0.000	12:10	
Control Test	0.076	12:10	
Air Blank	0.000	12:11	
Subject Sample	# 1	12:11	Subject Test Refused
Breath Volume	0.000	12:11	
Air Blank	0.000	12:12	
Control Test	0.080	12:12	
Air Blank	0.000	12:12	
Diagnostics Check	OK	12:13	

Data Download Date/Time: 12/23/2017 19:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JEUDY,RINALDO,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: MONCADASALAZAR,TERRY
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080AI
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:33

SUBJECT'S LAST NAME: DANIELS

DATE/TIME OF TEST: 11/16/2017 04:08:13

Diagnostics Check	OK	04:02
Air Blank	0.000	04:02
Control Test	0.082	04:02
Air Blank	0.000	04:03
Subject Sample # 1	0.000	04:03
Breath Volume	2.378	04:03
Air Blank	0.000	04:04
Air Blank	0.000	04:06
Subject Sample # 2	0.000	04:06
Breath Volume	2.808	04:06
Air Blank	0.000	04:07
Control Test	0.081	04:07
Air Blank	0.000	04:08
Diagnostics Check	OK	04:08

Data Download Date/Time: 12/23/2017 19:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GETSEE,JOY,S
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: KALLMEYER,DANIEL
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 11/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080AI
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:08

SUBJECT'S LAST NAME: GREGORY

DATE/TIME OF TEST: 11/17/2017 00:46:03

Diagnostics Check	OK	00:39
Air Blank	0.000	00:40
Control Test	0.080	00:40
Air Blank	0.000	00:41
Subject Sample # 1	0.145	00:41
Breath Volume	1.496	00:41
Air Blank	0.000	00:42
Air Blank	0.000	00:44
Subject Sample # 2	0.142	00:44
Breath Volume	2.062	00:44
Air Blank	0.000	00:45
Control Test	0.080	00:45
Air Blank	0.000	00:45
Diagnostics Check	OK	00:46

Data Download Date/Time: 12/23/2017 19:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAWES,ERIC,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: LAWES,ERIC
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080AI
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:52

SUBJECT'S LAST NAME: MULLINS

DATE/TIME OF TEST: 11/17/2017 20:26:49

Diagnostics Check	OK	20:19
Air Blank	0.000	20:19
Control Test	0.081	20:20
Air Blank	0.000	20:20
Subject Sample # 1	0.164	20:21
Breath Volume	3.261	20:21
Air Blank	0.000	20:21
Air Blank	0.000	20:23
Subject Sample # 2	0.162	20:25
Breath Volume	3.472	20:25
Air Blank	0.000	20:25
Control Test	0.079	20:26
Air Blank	0.000	20:26
Diagnostics Check	OK	20:26

Data Download Date/Time: 12/23/2017 19:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POTTER, ADAM, H
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: POTTER, ADAM
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080AI
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:19

SUBJECT'S LAST NAME: MACEDO

DATE/TIME OF TEST: 11/18/2017 02:52:48

Diagnostics Check	OK	02:45
Air Blank	0.000	02:46
Control Test	0.081	02:46
Air Blank	0.000	02:46
Subject Sample # 1	0.145	02:47
Breath Volume	2.308	02:47
Air Blank	0.000	02:48
Air Blank	0.000	02:50
Subject Sample # 2	0.146	02:51
Breath Volume	1.253	02:51
Air Blank	0.000	02:51
Control Test	0.080	02:52
Air Blank	0.000	02:52
Diagnostics Check	OK	02:52

Data Download Date/Time: 12/23/2017 19:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELANEY,AMY,L
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: CRUMMER,DAVID
ARRESTING OFFICER AGENCY: Winter Haven PD
LAST AGENCY INSPECTION: 11/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080AI
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 06:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:35

SUBJECT'S LAST NAME: ALDACO

DATE/TIME OF TEST: 11/19/2017 11:06:29

Diagnostics Check	OK	10:59
Air Blank	0.000	11:00
Control Test	0.081	11:00
Air Blank	0.000	11:01
Subject Sample # 1	0.031	11:01
Breath Volume	1.265	11:01
Air Blank	0.000	11:02
Air Blank	0.000	11:04
Subject Sample # 2	0.033	11:05
Breath Volume	1.605	11:05
Air Blank	0.000	11:05
Control Test	0.081	11:05
Air Blank	0.000	11:06
Diagnostics Check	OK	11:06

Data Download Date/Time: 12/23/2017 19:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHRISTELLO,CLIFFORD,A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: CHRISTELLO,CLIFFORD
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080AI
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:24

SUBJECT'S LAST NAME: MYERS

DATE/TIME OF TEST: 11/20/2017 20:50:34

Diagnostics Check	OK	20:44
Air Blank	0.000	20:45
Control Test	0.081	20:45
Air Blank	0.000	20:45
Subject Sample # 1	0.000	20:46
Breath Volume	1.843	20:46
Air Blank	0.000	20:46
Air Blank	0.000	20:48
Subject Sample # 2	0.000	20:49
Breath Volume	1.757	20:49
Air Blank	0.000	20:49
Control Test	0.081	20:49
Air Blank	0.000	20:50
Diagnostics Check	OK	20:50

Data Download Date/Time: 12/23/2017 19:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHRISTELLO, CLIFFORD, A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: WOFFORD, CLINT
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080AI
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: PATTERSON

DATE/TIME OF TEST: 11/20/2017 23:06:15

Diagnostics Check	OK	22:59
Air Blank	0.000	23:00
Control Test	0.080	23:00
Air Blank	0.000	23:01
Subject Sample # 1	0.133	23:01
Breath Volume	1.429	23:01
Air Blank	0.000	23:02
Air Blank	0.000	23:04
Subject Sample # 2	0.136	23:04
Breath Volume	2.011	23:04
Air Blank	0.000	23:05
Control Test	0.079	23:05
Air Blank	0.000	23:06
Diagnostics Check	OK	23:06

Data Download Date/Time: 12/23/2017 19:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LACAPRA,DANIEL,C
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: FENDER,KENNETH
ARRESTING OFFICER AGENCY: Bartow PD
LAST AGENCY INSPECTION: 11/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080AI
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 05:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:58

SUBJECT'S LAST NAME: HAYGOOD

DATE/TIME OF TEST: 11/21/2017 06:37:12

Diagnostics Check	OK	06:34	
Air Blank	0.000	06:34	
Control Test	0.081	06:35	
Air Blank	0.000	06:35	
Subject Sample	# 1	06:35	Subject Test Refused
Breath Volume	0.000	06:35	
Air Blank	0.000	06:36	
Control Test	0.080	06:36	
Air Blank	0.000	06:37	
Diagnostics Check	OK	06:37	

Data Download Date/Time: 12/23/2017 19:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUCAS,NATHANIEL,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: LUCAS,NATHANIEL
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080AI
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 18:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:42

SUBJECT'S LAST NAME: FOLEY

DATE/TIME OF TEST: 11/22/2017 19:19:08

Diagnostics Check	OK	19:13
Air Blank	0.000	19:13
Control Test	0.080	19:13
Air Blank	0.000	19:14
Subject Sample # 1	0.186	19:14
Breath Volume	1.460	19:14
Air Blank	0.000	19:15
Air Blank	0.000	19:17
Subject Sample # 2	0.184	19:17
Breath Volume	2.105	19:17
Air Blank	0.000	19:18
Control Test	0.080	19:18
Air Blank	0.000	19:19
Diagnostics Check	OK	19:19

Data Download Date/Time: 12/23/2017 19:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SINCLAIR,CAMERON,W
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: SINCLAIR,CAMERON
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080AI
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:37

SUBJECT'S LAST NAME: ROMAN JR.

DATE/TIME OF TEST: 11/22/2017 23:08:27

Diagnostics Check	OK	23:02
Air Blank	0.000	23:02
Control Test	0.081	23:03
Air Blank	0.000	23:03
Subject Sample # 1	0.305	23:04
Breath Volume	2.289	23:04
Air Blank	0.000	23:04
Air Blank	0.000	23:06
Subject Sample # 2	0.319	23:06
Breath Volume	1.574	23:06
Air Blank	0.000	23:07
Control Test	0.080	23:07
Air Blank	0.000	23:08
Diagnostics Check	OK	23:08

Data Download Date/Time: 12/23/2017 19:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANDERSON, DAVID, A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: MCKEE, JOHN
ARRESTING OFFICER AGENCY: Bartow PD
LAST AGENCY INSPECTION: 11/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080AI
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 14:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:17

SUBJECT'S LAST NAME: MCKELLIP

DATE/TIME OF TEST: 11/24/2017 15:45:15

Diagnostics Check	OK	15:38
Air Blank	0.000	15:38
Control Test	0.080	15:39
Air Blank	0.000	15:39
Subject Sample # 1	0.268	15:40
Breath Volume	1.746	15:40
Air Blank	0.000	15:41
Air Blank	0.000	15:43
Subject Sample # 2	0.277	15:43
Breath Volume	1.679	15:43
Air Blank	0.000	15:44
Control Test	0.080	15:44
Air Blank	0.000	15:45
Diagnostics Check	OK	15:45

Data Download Date/Time: 12/23/2017 19:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SINCLAIR,CAMERON,W
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: SINCLAIR,CAMERON
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080AI
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:06

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 11/25/2017 20:45:00

Diagnostics Check	OK	20:37
Air Blank	0.000	20:38
Control Test	0.080	20:38
Air Blank	0.000	20:39
Subject Sample # 1	0.231	20:40
Breath Volume	1.375	20:40
Air Blank	0.000	20:41
Air Blank	0.000	20:42
Subject Sample # 2	0.235	20:43
Breath Volume	1.406	20:43
Air Blank	0.000	20:44
Control Test	0.079	20:44
Air Blank	0.000	20:44
Diagnostics Check	OK	20:45

Data Download Date/Time: 12/23/2017 19:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUCAS,NATHANIEL,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: LUCAS,NATHANIEL
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080AI
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: GOFF

DATE/TIME OF TEST: 11/29/2017 22:40:03

Diagnostics Check	OK	22:33
Air Blank	0.000	22:34
Control Test	0.081	22:34
Air Blank	0.000	22:35
Subject Sample # 1	0.058	22:35
Breath Volume	1.949	22:35
Air Blank	0.000	22:36
Air Blank	0.000	22:38
Subject Sample # 2	0.055	22:38
Breath Volume	1.703	22:38
Air Blank	0.000	22:39
Control Test	0.080	22:39
Air Blank	0.000	22:39
Diagnostics Check	OK	22:40

Data Download Date/Time: 12/23/2017 19:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PHILLIPS,JORDAN,W
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: DURRANCE,JACOB
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080AI
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: PROB/PAROL
ARREST TIME: 02:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: WESTBROOK

DATE/TIME OF TEST: 11/30/2017 03:28:35

Diagnostics Check	OK	03:21
Air Blank	0.000	03:22
Control Test	0.080	03:22
Air Blank	0.000	03:23
Subject Sample # 1	0.070	03:23
Breath Volume	3.417	03:23
Air Blank	0.000	03:24
Air Blank	0.000	03:26
Subject Sample # 2	0.073	03:27
Breath Volume	3.398	03:27
Air Blank	0.000	03:27
Control Test	0.080	03:27
Air Blank	0.000	03:28
Diagnostics Check	OK	03:28

Data Download Date/Time: 12/23/2017 19:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001056

INSTRUMENT REGISTERED TO: DAVIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STRUNK,J,D
BREATH TEST OPERATOR AGENCY: Davie PD
ARRESTING OFFICER: PELTACK,J
ARRESTING OFFICER AGENCY: Davie PD
LAST AGENCY INSPECTION: 10/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG530001
DRY GAS EXPIRY DATE: 10/27/2017
VIOLATION CODE: DUI
ARREST TIME: 05:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:31

SUBJECT'S LAST NAME: COHN

DATE/TIME OF TEST: 11/05/2017 05:59:12

Diagnostics Check	OK	05:51
Air Blank	0.000	05:52
Control Test	0.081	05:52
Air Blank	0.000	05:53
Subject Sample # 1	0.133	05:53
Breath Volume	1.656	05:53
Air Blank	0.000	05:54
Air Blank	0.000	05:56
Subject Sample # 2	0.138	05:57
Breath Volume	1.792	05:57
Air Blank	0.000	05:58
Control Test	0.082	05:58
Air Blank	0.000	05:59
Diagnostics Check	OK	05:59

Data Download Date/Time: 11/28/2017 10:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001056

INSTRUMENT REGISTERED TO: DAVIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BART,STEVEN,G
BREATH TEST OPERATOR AGENCY: Davie PD
ARRESTING OFFICER: PENA,RYAN
ARRESTING OFFICER AGENCY: Davie PD
LAST AGENCY INSPECTION: 10/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG530001
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 03:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:12

SUBJECT'S LAST NAME: LONDONO

DATE/TIME OF TEST: 11/11/2017 04:40:53

Diagnostics Check	OK	04:32
Air Blank	0.000	04:32
Control Test	0.081	04:33
Air Blank	0.000	04:33
Subject Sample # 1	0.157	04:35
Breath Volume	1.632	04:35
Air Blank	0.000	04:36
Air Blank	0.000	04:38
Subject Sample # 2	0.167	04:39
Breath Volume	1.312	04:39
Air Blank	0.000	04:39
Control Test	0.079	04:40
Air Blank	0.000	04:40
Diagnostics Check	OK	04:40

Data Download Date/Time: 11/28/2017 10:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001056

INSTRUMENT REGISTERED TO: DAVIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BART,STEVEN,G
BREATH TEST OPERATOR AGENCY: Davie PD
ARRESTING OFFICER: PELTACK,JOHN
ARRESTING OFFICER AGENCY: Davie PD
LAST AGENCY INSPECTION: 10/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG530001
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 05:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:56

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 11/25/2017 06:25:41

Diagnostics Check	OK	06:18
Air Blank	0.000	06:19
Control Test	0.080	06:19
Air Blank	0.000	06:19
Subject Sample # 1	0.226	06:20
Breath Volume	2.148	06:20
Air Blank	0.000	06:21
Air Blank	0.000	06:22
Subject Sample # 2	0.238	06:24
Breath Volume	2.753	06:24
Air Blank	0.000	06:24
Control Test	0.080	06:25
Air Blank	0.000	06:25
Diagnostics Check	OK	06:25

Data Download Date/Time: 11/28/2017 10:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001060

INSTRUMENT REGISTERED TO: WINTER HAVEN PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NASH,BRIAN,T
BREATH TEST OPERATOR AGENCY: Winter Haven PD
ARRESTING OFFICER: BOONE,BRANDON
ARRESTING OFFICER AGENCY: Winter Haven PD
LAST AGENCY INSPECTION: 11/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 17:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:35

SUBJECT'S LAST NAME: CORTES

DATE/TIME OF TEST: 11/30/2017 18:02:40

Diagnostics Check	OK	17:55
Air Blank	0.000	17:56
Control Test	0.081	17:56
Air Blank	0.000	17:56
Subject Sample # 1	0.293	17:57
Breath Volume	1.941	17:57
Air Blank	0.000	17:58
Air Blank	0.000	18:00
Subject Sample # 2	0.278	18:00
Breath Volume	2.359	18:00
Air Blank	0.000	18:01
Control Test	0.081	18:02
Air Blank	0.000	18:02
Diagnostics Check	OK	18:02

Data Download Date/Time: 12/12/2017 12:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001061

INSTRUMENT REGISTERED TO: GULFPORT PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: CARTER,TIMOTHY,R

BREATH TEST OPERATOR AGENCY: Gulfport PD

ARRESTING OFFICER: SMITH,CORY

ARRESTING OFFICER AGENCY: Gulfport PD

LAST AGENCY INSPECTION: 10/31/2017

Date or Time Changed: No

DRY GAS STANDARD LOT #: 17817080A2

DRY GAS EXPIRY DATE: 08/05/2019

VIOLATION CODE: DUI

ARREST TIME: 05:14

GENDER: Male

OBSERVATION PERIOD BEGAN: 05:46

SUBJECT'S LAST NAME: CARLSTROM

DATE/TIME OF TEST: 11/03/2017 06:14:03

Diagnostics Check	OK	06:07
Air Blank	0.000	06:07
Control Test	0.081	06:07
Air Blank	0.000	06:08
Subject Sample # 1	0.000	06:09
Breath Volume	2.179	06:09
Air Blank	0.000	06:09
Air Blank	0.000	06:11
Subject Sample # 2	0.000	06:12
Breath Volume	3.187	06:12
Air Blank	0.000	06:13
Control Test	0.081	06:13
Air Blank	0.000	06:13
Diagnostics Check	OK	06:14

Data Download Date/Time: 11/26/2017 04:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001061

INSTRUMENT REGISTERED TO: GULFPORT PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILTSE,EVA,
BREATH TEST OPERATOR AGENCY: Gulfport PD
ARRESTING OFFICER: PALAZZOLO,COURTNEY
ARRESTING OFFICER AGENCY: Gulfport PD
LAST AGENCY INSPECTION: 11/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 19:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:30

SUBJECT'S LAST NAME: HANNULA

DATE/TIME OF TEST: 11/29/2017 19:56:48

Diagnostics Check	OK	19:50
Air Blank	0.000	19:51
Control Test	0.080	19:51
Air Blank	0.000	19:52
Subject Sample # 1	0.189	19:52
Breath Volume	1.343	19:52
Air Blank	0.000	19:53
Air Blank	0.000	19:54
Subject Sample # 2	0.191	19:55
Breath Volume	1.101	19:55
Air Blank	0.000	19:55
Control Test	0.080	19:56
Air Blank	0.000	19:56
Diagnostics Check	OK	19:56

Data Download Date/Time: 12/31/2017 15:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001062

INSTRUMENT REGISTERED TO: TREASURE ISLAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SWETNICH,JARED,A
BREATH TEST OPERATOR AGENCY: Treasure Island PD
ARRESTING OFFICER: JANIK,STEPHEN
ARRESTING OFFICER AGENCY: Indian Shores PD
LAST AGENCY INSPECTION: 10/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A6
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:09

SUBJECT'S LAST NAME: FINCANNON

DATE/TIME OF TEST: 11/04/2017 03:37:26

Diagnostics Check	OK	03:29
Air Blank	0.000	03:30
Control Test	0.080	03:30
Air Blank	0.000	03:31
Subject Sample # 1	0.168	03:32
Breath Volume	1.972	03:32
Air Blank	0.000	03:33
Air Blank	0.000	03:35
Subject Sample # 2	0.170	03:35
Breath Volume	1.445	03:35
Air Blank	0.000	03:36
Control Test	0.079	03:36
Air Blank	0.000	03:37
Diagnostics Check	OK	03:37

Data Download Date/Time: 11/22/2017 09:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001062

INSTRUMENT REGISTERED TO: TREASURE ISLAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SWETNICH,JARED,A
BREATH TEST OPERATOR AGENCY: Treasure Island PD
ARRESTING OFFICER: JANIK,STEPHEN
ARRESTING OFFICER AGENCY: Indian Shores PD
LAST AGENCY INSPECTION: 10/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A6
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:37

SUBJECT'S LAST NAME: BUDESA

DATE/TIME OF TEST: 11/10/2017 04:04:47

Diagnostics Check	OK	04:01	
Air Blank	0.000	04:02	
Control Test	0.079	04:02	
Air Blank	0.000	04:03	
Subject Sample	# 1	04:03	Subject Test Refused
Breath Volume	0.000	04:03	
Air Blank	0.000	04:03	
Control Test	0.080	04:04	
Air Blank	0.000	04:04	
Diagnostics Check	OK	04:04	

Data Download Date/Time: 11/22/2017 09:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001063

INSTRUMENT REGISTERED TO: HOLLYWOOD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BALUJA,JONATHAN,
BREATH TEST OPERATOR AGENCY: Hollywood PD
ARRESTING OFFICER: BALUJA,JONATHAN
ARRESTING OFFICER AGENCY: Hollywood PD
LAST AGENCY INSPECTION: 10/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 20:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: ALBERTARIO

DATE/TIME OF TEST: 11/03/2017 21:33:28

Diagnostics Check	OK	21:25
Air Blank	0.000	21:25
Control Test	0.077	21:26
Air Blank	0.000	21:26
Subject Sample # 1	0.168	21:27
Breath Volume	2.312	21:27
Air Blank	0.000	21:28
Air Blank	0.000	21:30
Subject Sample # 2	0.168	21:31
Breath Volume	2.558	21:31
Air Blank	0.000	21:32
Control Test	0.076	21:32
Air Blank	0.000	21:33
Diagnostics Check	OK	21:33

Data Download Date/Time: 11/24/2017 14:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001063

INSTRUMENT REGISTERED TO: HOLLYWOOD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,MELANIE,A
BREATH TEST OPERATOR AGENCY: Hollywood PD
ARRESTING OFFICER: GARCIA,MELANIE
ARRESTING OFFICER AGENCY: Hollywood PD
LAST AGENCY INSPECTION: 11/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 18:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:36

SUBJECT'S LAST NAME: GOMEZ PEREZ

DATE/TIME OF TEST: 11/26/2017 19:05:08

Diagnostics Check	OK	18:58
Air Blank	0.000	18:58
Control Test	0.077	18:59
Air Blank	0.000	18:59
Subject Sample # 1	0.215	19:00
Breath Volume	2.781	19:00
Air Blank	0.000	19:01
Air Blank	0.000	19:03
Subject Sample # 2	0.217	19:03
Breath Volume	2.175	19:03
Air Blank	0.000	19:04
Control Test	0.075	19:04
Air Blank	0.000	19:05
Diagnostics Check	OK	19:05

Data Download Date/Time: 12/03/2017 11:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

INSTRUMENT REGISTERED TO: NEW PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,ROBERT,R
BREATH TEST OPERATOR AGENCY: Port Richey PD
ARRESTING OFFICER: WILLIAMS,ROBERT
ARRESTING OFFICER AGENCY: Port Richey PD
LAST AGENCY INSPECTION: 10/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:07

SUBJECT'S LAST NAME: ZAPPASODI

DATE/TIME OF TEST: 11/09/2017 23:35:57

Diagnostics Check	OK	23:27
Air Blank	0.000	23:28
Control Test	0.079	23:28
Air Blank	0.000	23:29
Subject Sample # 1	0.000	23:31
Breath Volume	1.132	23:31
Air Blank	0.000	23:32
Air Blank	0.000	23:34
Subject Sample # 2	0.000	23:34
Breath Volume	1.503	23:34
Air Blank	0.000	23:35
Control Test	0.079	23:35
Air Blank	0.000	23:35
Diagnostics Check	OK	23:35

Data Download Date/Time: 11/28/2017 22:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

INSTRUMENT REGISTERED TO: NEW PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRAPNELL,CHRISTOPHER,P
BREATH TEST OPERATOR AGENCY: New Port Richey PD
ARRESTING OFFICER: MILANA,WILLIAM
ARRESTING OFFICER AGENCY: New Port Richey PD
LAST AGENCY INSPECTION: 10/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:56

SUBJECT'S LAST NAME: TESTA

DATE/TIME OF TEST: 11/26/2017 03:25:01

Diagnostics Check	OK	03:18
Air Blank	0.000	03:19
Control Test	0.078	03:19
Air Blank	0.000	03:20
Subject Sample # 1	0.169	03:20
Breath Volume	3.542	03:20
Air Blank	0.000	03:21
Air Blank	0.000	03:22
Subject Sample # 2	0.169	03:23
Breath Volume	3.316	03:23
Air Blank	0.000	03:24
Control Test	0.078	03:24
Air Blank	0.000	03:24
Diagnostics Check	OK	03:25

Data Download Date/Time: 11/28/2017 22:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001065

INSTRUMENT REGISTERED TO: TARPON SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRINIDAD,MICHAEL,E
BREATH TEST OPERATOR AGENCY: Tarpon Springs PD
ARRESTING OFFICER: TRINIDAD,MICHAEL
ARRESTING OFFICER AGENCY: Tarpon Springs PD
LAST AGENCY INSPECTION: 10/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A3
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 16:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:14

SUBJECT'S LAST NAME: KARSHNESKI

DATE/TIME OF TEST: 11/02/2017 17:41:06

Diagnostics Check	OK	17:34
Air Blank	0.000	17:35
Control Test	0.079	17:35
Air Blank	0.000	17:35
Subject Sample # 1	0.181	17:36
Breath Volume	1.503	17:36
Air Blank	0.000	17:37
Air Blank	0.000	17:39
Subject Sample # 2	0.180	17:39
Breath Volume	1.996	17:39
Air Blank	0.000	17:40
Control Test	0.078	17:40
Air Blank	0.000	17:41
Diagnostics Check	OK	17:41

Data Download Date/Time: 11/27/2017 13:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001065

INSTRUMENT REGISTERED TO: TARPON SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPATZ,JOHN,M
BREATH TEST OPERATOR AGENCY: Tarpon Springs PD
ARRESTING OFFICER: SPARKS,JOSHUA
ARRESTING OFFICER AGENCY: Tarpon Springs PD
LAST AGENCY INSPECTION: 10/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A3
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: SCHINELLI

DATE/TIME OF TEST: 11/04/2017 02:33:49

Diagnostics Check	OK	02:31	
Air Blank	0.000	02:31	
Control Test	0.079	02:31	
Air Blank	0.000	02:32	
Subject Sample	# 1	02:32	Subject Test Refused
Breath Volume	0.000	02:32	
Air Blank	0.000	02:32	
Control Test	0.079	02:33	
Air Blank	0.000	02:33	
Diagnostics Check	OK	02:33	

Data Download Date/Time: 11/27/2017 13:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001065

INSTRUMENT REGISTERED TO: TARPON SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRINIDAD, MICHAEL, E
BREATH TEST OPERATOR AGENCY: Tarpon Springs PD
ARRESTING OFFICER: BRINKER, C
ARRESTING OFFICER AGENCY: Tarpon Springs PD
LAST AGENCY INSPECTION: 10/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A3
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:18

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 11/24/2017 21:43:31

Diagnostics Check	OK	21:39	
Air Blank	0.000	21:39	
Control Test	0.078	21:39	
Air Blank	0.000	21:40	
Subject Sample	# 1	21:42	Subject Test Refused
Breath Volume	0.000	21:42	
Air Blank	0.000	21:42	
Control Test	0.078	21:42	
Air Blank	0.000	21:43	
Diagnostics Check	OK	21:43	

Data Download Date/Time: 11/27/2017 13:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KASELAK,STEPHEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 23:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:03

SUBJECT'S LAST NAME: CARPENTIER

DATE/TIME OF TEST: 11/10/2017 00:26:40

Diagnostics Check	OK	00:23	
Air Blank	0.000	00:24	
Control Test	0.080	00:24	
Air Blank	0.000	00:25	
Subject Sample	# 1	00:25	Subject Test Refused
Breath Volume	0.000	00:25	
Air Blank	0.000	00:25	
Control Test	0.080	00:26	
Air Blank	0.000	00:26	
Diagnostics Check	OK	00:26	

Data Download Date/Time: 11/29/2017 20:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HORNING,PETER,T
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: CAREY,KEVIN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 08/24/2018
VIOLATION CODE: DUI
ARREST TIME: 11:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 12:38

SUBJECT'S LAST NAME: WEBB

DATE/TIME OF TEST: 11/10/2017 13:05:29

Diagnostics Check	OK	12:59
Air Blank	0.000	12:59
Control Test	0.079	13:00
Air Blank	0.000	13:00
Subject Sample # 1	0.227	13:01
Breath Volume	2.105	13:01
Air Blank	0.000	13:01
Air Blank	0.000	13:03
Subject Sample # 2	0.226	13:03
Breath Volume	1.585	13:03
Air Blank	0.000	13:04
Control Test	0.078	13:04
Air Blank	0.000	13:05
Diagnostics Check	OK	13:05

Data Download Date/Time: 11/29/2017 20:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KASELAK,STEPHEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: MINTON

DATE/TIME OF TEST: 11/11/2017 01:06:04

Diagnostics Check	OK	01:03	
Air Blank	0.000	01:03	
Control Test	0.080	01:03	
Air Blank	0.000	01:04	
Subject Sample	# 1	01:04	Subject Test Refused
Breath Volume	0.000	01:04	
Air Blank	0.000	01:05	
Control Test	0.079	01:05	
Air Blank	0.000	01:05	
Diagnostics Check	OK	01:06	

Data Download Date/Time: 11/29/2017 20:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ENGLAND,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: ENGLAND,CHRISTOPHER
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 10/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:28

SUBJECT'S LAST NAME: KNOWLTON

DATE/TIME OF TEST: 11/03/2017 22:55:55

Diagnostics Check	OK	22:48
Air Blank	0.000	22:49
Control Test	0.080	22:49
Air Blank	0.000	22:50
Subject Sample # 1	0.164	22:51
Breath Volume	2.285	22:51
Air Blank	0.000	22:51
Air Blank	0.000	22:53
Subject Sample # 2	0.164	22:54
Breath Volume	1.656	22:54
Air Blank	0.000	22:54
Control Test	0.079	22:55
Air Blank	0.000	22:55
Diagnostics Check	OK	22:55

Data Download Date/Time: 11/15/2017 02:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HATCH,M,A
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: HATCH,M
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 10/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 05:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:35

SUBJECT'S LAST NAME: SAMPSELL

DATE/TIME OF TEST: 11/04/2017 06:04:25

Diagnostics Check	OK	05:57
Air Blank	0.000	05:58
Control Test	0.080	05:58
Air Blank	0.000	05:58
Subject Sample # 1	0.182	05:59
Breath Volume	2.375	05:59
Air Blank	0.000	06:00
Air Blank	0.000	06:02
Subject Sample # 2	0.180	06:02
Breath Volume	2.812	06:02
Air Blank	0.000	06:03
Control Test	0.078	06:03
Air Blank	0.000	06:04
Diagnostics Check	OK	06:04

Data Download Date/Time: 11/15/2017 02:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NUGENT,DAVID,A
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: HURT,JONATHAN
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 10/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 11/05/2017 01:13:40

Diagnostics Check	OK	01:07
Air Blank	0.000	01:07
Control Test	0.080	01:07
Air Blank	0.000	01:08
Subject Sample # 1	0.000	01:09
Breath Volume	1.648	01:09
Air Blank	0.000	01:09
Air Blank	0.000	01:11
Subject Sample # 2	0.000	01:12
Breath Volume	2.332	01:12
Air Blank	0.000	01:12
Control Test	0.079	01:13
Air Blank	0.000	01:13
Diagnostics Check	OK	01:13

Data Download Date/Time: 11/15/2017 02:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NEGERSMITH,D,M
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: NEGERSMITH,D
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 10/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: BE BOUT

DATE/TIME OF TEST: 11/05/2017 02:30:27

Diagnostics Check	OK	02:22
Air Blank	0.000	02:23
Control Test	0.079	02:23
Air Blank	0.000	02:23
Subject Sample # 1	0.000	02:25
Breath Volume	1.347	02:25
Air Blank	0.000	02:26
Air Blank	0.000	02:28
Subject Sample # 2	0.000	02:29
Breath Volume	1.492	02:29
Air Blank	0.000	02:29
Control Test	0.078	02:29
Air Blank	0.000	02:30
Diagnostics Check	OK	02:30

Data Download Date/Time: 11/15/2017 02:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NUGENT,DAVID,A
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: NUGENT,DAVID
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 10/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: CALIHUA PANZO

DATE/TIME OF TEST: 11/08/2017 23:31:44

Diagnostics Check	OK	23:20	
Air Blank	0.000	23:20	
Control Test	0.080	23:21	
Air Blank	0.000	23:21	
Subject Sample	# 1	23:24	No Sample Provided
Breath Volume	0.000	23:24	
Air Blank	0.000	23:25	
Air Blank	0.000	23:27	
Subject Sample	# 2	23:30	No Sample Provided
Breath Volume	0.000	23:30	
Air Blank	0.000	23:30	
Control Test	0.079	23:31	
Air Blank	0.000	23:31	
Diagnostics Check	OK	23:31	

Data Download Date/Time: 11/15/2017 02:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NUGENT,DAVID,A
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: NUGENT,DAVID
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 10/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:25

SUBJECT'S LAST NAME: BYRUM

DATE/TIME OF TEST: 11/10/2017 20:53:18

Diagnostics Check	OK	20:47
Air Blank	0.000	20:47
Control Test	0.080	20:47
Air Blank	0.000	20:48
Subject Sample # 1	0.000	20:48
Breath Volume	2.144	20:48
Air Blank	0.000	20:49
Air Blank	0.000	20:51
Subject Sample # 2	0.000	20:51
Breath Volume	2.261	20:51
Air Blank	0.000	20:52
Control Test	0.080	20:52
Air Blank	0.000	20:53
Diagnostics Check	OK	20:53

Data Download Date/Time: 11/15/2017 02:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NUGENT,DAVID,A
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: NUGENT,DAVID
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 10/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:17

SUBJECT'S LAST NAME: GRAY

DATE/TIME OF TEST: 11/11/2017 19:51:39

Diagnostics Check	OK	19:40	
Air Blank	0.000	19:40	
Control Test	0.080	19:40	
Air Blank	0.000	19:41	
Subject Sample	# 1	19:44	No Sample Provided
Breath Volume	0.000	19:44	
Air Blank	0.000	19:45	
Air Blank	0.000	19:47	
Subject Sample	# 2	19:50	No Sample Provided
Breath Volume	0.000	19:50	
Air Blank	0.000	19:50	
Control Test	0.080	19:51	
Air Blank	0.000	19:51	
Diagnostics Check	OK	19:51	

Data Download Date/Time: 11/15/2017 02:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REED,CHAD,R
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: REED,CHAD
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 11/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: MOYNAHAN

DATE/TIME OF TEST: 11/16/2017 03:13:09

Diagnostics Check	OK	03:01	
Air Blank	0.000	03:02	
Control Test	0.080	03:02	
Air Blank	0.000	03:03	
Subject Sample	# 1	03:06	No Sample Provided
Breath Volume	0.000	03:06	
Air Blank	0.000	03:06	
Air Blank	0.000	03:08	
Subject Sample	# 2	03:11	No Sample Provided
Breath Volume	0.000	03:11	
Air Blank	0.000	03:12	
Control Test	0.080	03:12	
Air Blank	0.000	03:13	
Diagnostics Check	OK	03:13	

Data Download Date/Time: 12/21/2017 20:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REED,CHAD,R
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: HURT,JONATHAN
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 11/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:21

SUBJECT'S LAST NAME: WHITTAKER

DATE/TIME OF TEST: 11/17/2017 00:50:21

Diagnostics Check	OK	00:41
Air Blank	0.000	00:41
Control Test	0.080	00:42
Air Blank	0.000	00:42
Subject Sample # 1	0.139	00:44
Breath Volume	2.140	00:44
Air Blank	0.000	00:46
Air Blank	0.000	00:47
Subject Sample # 2	0.135	00:48
Breath Volume	2.183	00:48
Air Blank	0.000	00:49
Control Test	0.076	00:49
Air Blank	0.000	00:50
Diagnostics Check	OK	00:50

Data Download Date/Time: 12/21/2017 20:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REED,CHAD,R
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: NEGERSMITH,DANIEL
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 11/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:38

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 11/17/2017 01:15:11

Diagnostics Check	OK	01:06
Air Blank	0.000	01:06
Control Test	0.077	01:07
Air Blank	0.000	01:07
Subject Sample # 1	0.049	01:10
Breath Volume	1.972	01:10
Air Blank	0.000	01:11
Air Blank	0.000	01:12
Subject Sample # 2	0.041	01:13
Breath Volume	1.703	01:13
Air Blank	0.000	01:14
Control Test	0.076	01:14
Air Blank	0.000	01:15
Diagnostics Check	OK	01:15

Data Download Date/Time: 12/21/2017 20:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REED,CHAD,R
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: NUGENT,DAVID
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 11/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:31

SUBJECT'S LAST NAME: WEAVER

DATE/TIME OF TEST: 11/17/2017 23:04:47

Diagnostics Check	OK	22:51	
Air Blank	0.000	22:52	
Control Test	0.079	22:52	
Air Blank	0.000	22:53	
Subject Sample	# 1 0.202	22:56	
Breath Volume	2.183	22:56	
Air Blank	0.000	22:57	
Air Blank	0.000	22:58	
Subject Sample	# 2 0.181	23:00	
Breath Volume	1.171	23:00	
Air Blank	0.000	23:01	
Air Blank	0.000	23:02	
Subject Sample	# 3 0.193	23:03	
Breath Volume	2.085	23:03	
Air Blank		23:04	Purge Fail
Air Blank	0.000	23:04	

Data Download Date/Time: 12/21/2017 20:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REED,CHAD,R
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: NUGENT,DAVID
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 11/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:31

SUBJECT'S LAST NAME: WEAVER

DATE/TIME OF TEST: 11/17/2017 23:08:56

Diagnostics Check	OK	23:07	
Air Blank	0.000	23:07	
Control Test	0.073	23:08	Control Outside Tolerance
Air Blank	0.000	23:08	

Data Download Date/Time: 12/21/2017 20:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REED,CHAD,R
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: NUGENT,DAVID
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 11/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:31

SUBJECT'S LAST NAME: WEAVER

DATE/TIME OF TEST: 11/17/2017 23:13:50

Diagnostics Check	OK	23:10	
Air Blank	0.000	23:11	
Control Test	0.077	23:11	
Air Blank	0.000	23:12	
Subject Sample # 1	0.193	23:12	
Breath Volume	2.000	23:12	
Air Blank		23:13	Purge Fail
Air Blank	0.000	23:13	

Data Download Date/Time: 12/21/2017 20:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REED,CHAD,R
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: NUGENT,DAVID
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 11/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:31

SUBJECT'S LAST NAME: WEAVER

DATE/TIME OF TEST: 11/17/2017 23:23:24

Diagnostics Check	OK	23:15
Air Blank	0.000	23:16
Control Test	0.077	23:16
Air Blank	0.000	23:17
Subject Sample # 1	0.183	23:18
Breath Volume	1.472	23:18
Air Blank	0.000	23:18
Air Blank	0.000	23:20
Subject Sample # 2	0.192	23:21
Breath Volume	1.449	23:21
Air Blank	0.000	23:22
Control Test	0.077	23:22
Air Blank	0.000	23:23
Diagnostics Check	OK	23:23

Data Download Date/Time: 12/21/2017 20:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REED,CHAD,R
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: NEGERSMITH,DANIEL
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 11/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 11/19/2017 02:57:20

Diagnostics Check	OK	02:50
Air Blank	0.000	02:50
Control Test	0.079	02:51
Air Blank	0.000	02:51
Subject Sample # 1	0.244	02:52
Breath Volume	1.777	02:52
Air Blank	0.000	02:53
Air Blank	0.000	02:54
Subject Sample # 2	0.237	02:55
Breath Volume	2.503	02:55
Air Blank	0.000	02:56
Control Test	0.077	02:56
Air Blank	0.000	02:57
Diagnostics Check	OK	02:57

Data Download Date/Time: 12/21/2017 20:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSEPILER,JONATHAN,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CARLSON,P
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 10/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:14

SUBJECT'S LAST NAME: MERRYMAN

DATE/TIME OF TEST: 11/01/2017 01:59:08

Diagnostics Check	OK	01:51	
Air Blank	0.000	01:51	
Control Test	0.081	01:51	
Air Blank	0.000	01:52	
Subject Sample	# 1 0.192	01:52	
Breath Volume	1.199	01:52	
Air Blank	0.000	01:53	
Air Blank	0.000	01:55	
Subject Sample	# 2 0.172	01:57	
Breath Volume	2.261	01:57	
Air Blank		01:58	Purge Fail
Air Blank		01:59	Purge Fail

Data Download Date/Time: 11/01/2017 11:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSEPILER,JONATHAN,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CARLSON,P
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 10/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:14

SUBJECT'S LAST NAME: MERRYMAN

DATE/TIME OF TEST: 11/01/2017 02:14:19

Diagnostics Check	OK	02:12	
Air Blank	0.000	02:13	
Control Test	0.069	02:13	Control Outside Tolerance
Air Blank	0.000	02:14	

Data Download Date/Time: 11/01/2017 11:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSEPILER,JONATHAN,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: TORRES,W
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 10/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:47

SUBJECT'S LAST NAME: CEBECK

DATE/TIME OF TEST: 11/01/2017 04:10:40

Diagnostics Check	OK	04:07	
Air Blank	0.000	04:08	
Control Test	0.077	04:08	
Air Blank	0.000	04:09	
Subject Sample	# 1	04:09	Subject Test Refused
Breath Volume	0.000	04:09	
Air Blank	0.000	04:09	
Control Test	0.077	04:10	
Air Blank	0.000	04:10	
Diagnostics Check	OK	04:10	

Data Download Date/Time: 11/01/2017 11:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARRIER,A,R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BRYANT,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:28

SUBJECT'S LAST NAME: GLEE

DATE/TIME OF TEST: 11/02/2017 00:55:15

Diagnostics Check	OK	00:48
Air Blank	0.000	00:49
Control Test	0.081	00:49
Air Blank	0.000	00:50
Subject Sample # 1	0.202	00:50
Breath Volume	2.285	00:50
Air Blank	0.000	00:51
Air Blank	0.000	00:53
Subject Sample # 2	0.207	00:53
Breath Volume	2.011	00:53
Air Blank	0.000	00:54
Control Test	0.081	00:54
Air Blank	0.000	00:55
Diagnostics Check	OK	00:55

Data Download Date/Time: 11/17/2017 13:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARRIER,A,R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CARLSON,P
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 11/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:26

SUBJECT'S LAST NAME: STRAUSSER

DATE/TIME OF TEST: 11/02/2017 05:52:39

Diagnostics Check	OK	05:46
Air Blank	0.000	05:46
Control Test	0.081	05:46
Air Blank	0.000	05:47
Subject Sample # 1	0.149	05:48
Breath Volume	2.902	05:48
Air Blank	0.000	05:48
Air Blank	0.000	05:50
Subject Sample # 2	0.154	05:51
Breath Volume	2.324	05:51
Air Blank	0.000	05:51
Control Test	0.080	05:52
Air Blank	0.000	05:52
Diagnostics Check	OK	05:52

Data Download Date/Time: 11/17/2017 13:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARRIER,A,R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MOELLER,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:17

SUBJECT'S LAST NAME: MCDOWELL

DATE/TIME OF TEST: 11/03/2017 02:41:35

Diagnostics Check	OK	02:38	
Air Blank	0.000	02:39	
Control Test	0.081	02:39	
Air Blank	0.000	02:40	
Subject Sample	# 1	02:40	Subject Test Refused
Breath Volume	0.000	02:40	
Air Blank	0.000	02:40	
Control Test	0.081	02:40	
Air Blank	0.000	02:41	
Diagnostics Check	OK	02:41	

Data Download Date/Time: 11/17/2017 13:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARRIER,A,R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHITE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:09

SUBJECT'S LAST NAME: SHULER

DATE/TIME OF TEST: 11/03/2017 03:39:19

Diagnostics Check	OK	03:30
Air Blank	0.000	03:30
Control Test	0.081	03:30
Air Blank	0.000	03:31
Subject Sample # 1	0.188	03:31
Breath Volume	2.011	03:31
Air Blank	0.000	03:32
Air Blank	0.000	03:34
Subject Sample # 2	0.161	03:34
Breath Volume	2.503	03:34
Air Blank	0.000	03:35
Air Blank	0.000	03:37
Subject Sample # 3	0.165	03:37
Breath Volume	3.027	03:37
Air Blank	0.000	03:38
Control Test	0.080	03:38
Air Blank	0.000	03:39
Diagnostics Check	OK	03:39

Data Download Date/Time: 11/17/2017 13:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBER,JOSEPH,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SCOTT,A.G.
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 05:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:45

SUBJECT'S LAST NAME: COSBY

DATE/TIME OF TEST: 11/05/2017 06:09:15

Diagnostics Check	OK	06:06	
Air Blank	0.000	06:06	
Control Test	0.081	06:06	
Air Blank	0.000	06:07	
Subject Sample	# 1	06:07	Subject Test Refused
Breath Volume	0.000	06:07	
Air Blank	0.000	06:08	
Control Test	0.081	06:08	
Air Blank	0.000	06:09	
Diagnostics Check	OK	06:09	

Data Download Date/Time: 11/17/2017 13:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARRIER,A,R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MOELLER,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:09

SUBJECT'S LAST NAME: BLAND

DATE/TIME OF TEST: 11/06/2017 00:36:47

Diagnostics Check	OK	00:30
Air Blank	0.000	00:30
Control Test	0.080	00:31
Air Blank	0.000	00:31
Subject Sample # 1	0.174	00:32
Breath Volume	2.898	00:32
Air Blank	0.000	00:32
Air Blank	0.000	00:34
Subject Sample # 2	0.169	00:35
Breath Volume	1.214	00:35
Air Blank	0.000	00:35
Control Test	0.080	00:36
Air Blank	0.000	00:36
Diagnostics Check	OK	00:36

Data Download Date/Time: 11/17/2017 13:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARRIER,A,R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DOUGHFMAN,D
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:54

SUBJECT'S LAST NAME: YOUNG

DATE/TIME OF TEST: 11/06/2017 01:17:17

Diagnostics Check	OK	01:14	
Air Blank	0.000	01:14	
Control Test	0.081	01:15	
Air Blank	0.000	01:15	
Subject Sample	# 1	01:15	Subject Test Refused
Breath Volume	0.000	01:15	
Air Blank	0.000	01:16	
Control Test	0.080	01:16	
Air Blank	0.000	01:17	
Diagnostics Check	OK	01:17	

Data Download Date/Time: 11/17/2017 13:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARRIER,A,R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FISETTE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:23

SUBJECT'S LAST NAME: LUKOWSKY

DATE/TIME OF TEST: 11/06/2017 04:47:06

Diagnostics Check	OK	04:44	
Air Blank	0.000	04:44	
Control Test	0.080	04:44	
Air Blank	0.000	04:45	
Subject Sample	# 1	04:45	Subject Test Refused
Breath Volume	0.000	04:45	
Air Blank	0.000	04:46	
Control Test	0.081	04:46	
Air Blank	0.000	04:47	
Diagnostics Check	OK	04:47	

Data Download Date/Time: 11/17/2017 13:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARRIER,A,R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FISETTE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:52

SUBJECT'S LAST NAME: WALKER

DATE/TIME OF TEST: 11/06/2017 23:19:10

Diagnostics Check	OK	23:12
Air Blank	0.000	23:13
Control Test	0.080	23:13
Air Blank	0.000	23:14
Subject Sample # 1	0.000	23:14
Breath Volume	1.542	23:14
Air Blank	0.000	23:15
Air Blank	0.000	23:17
Subject Sample # 2	0.000	23:17
Breath Volume	2.070	23:17
Air Blank	0.000	23:18
Control Test	0.081	23:18
Air Blank	0.000	23:19
Diagnostics Check	OK	23:19

Data Download Date/Time: 11/17/2017 13:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARRIER,A,R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: STANLEY,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:26

SUBJECT'S LAST NAME: THIEL

DATE/TIME OF TEST: 11/07/2017 01:53:12

Diagnostics Check	OK	01:46
Air Blank	0.000	01:46
Control Test	0.080	01:47
Air Blank	0.000	01:47
Subject Sample # 1	0.107	01:48
Breath Volume	2.371	01:48
Air Blank	0.000	01:49
Air Blank	0.000	01:51
Subject Sample # 2	0.107	01:51
Breath Volume	1.480	01:51
Air Blank	0.000	01:52
Control Test	0.079	01:52
Air Blank	0.000	01:53
Diagnostics Check	OK	01:53

Data Download Date/Time: 11/17/2017 13:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARRIER,A,R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DOUGHFAM,D
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:28

SUBJECT'S LAST NAME: TIMMONS

DATE/TIME OF TEST: 11/07/2017 04:55:14

Diagnostics Check	OK	04:48
Air Blank	0.000	04:49
Control Test	0.080	04:49
Air Blank	0.000	04:49
Subject Sample # 1	0.148	04:50
Breath Volume	1.492	04:50
Air Blank	0.000	04:51
Air Blank	0.000	04:53
Subject Sample # 2	0.142	04:53
Breath Volume	1.277	04:53
Air Blank	0.000	04:54
Control Test	0.080	04:54
Air Blank	0.000	04:55
Diagnostics Check	OK	04:55

Data Download Date/Time: 11/17/2017 13:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARRIER,A,R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: GARCIA,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:04

SUBJECT'S LAST NAME: FOLEY

DATE/TIME OF TEST: 11/14/2017 02:03:39

Diagnostics Check	OK	02:00	
Air Blank	0.000	02:01	
Control Test	0.081	02:01	
Air Blank	0.000	02:02	
Subject Sample	# 1	02:02	Subject Test Refused
Breath Volume	0.000	02:02	
Air Blank	0.000	02:02	
Control Test	0.081	02:03	
Air Blank	0.000	02:03	
Diagnostics Check	OK	02:03	

Data Download Date/Time: 11/17/2017 13:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARRIER,A,R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: THOMPSON,M
ARRESTING OFFICER AGENCY: Atlantic Beach PD
LAST AGENCY INSPECTION: 11/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:22

SUBJECT'S LAST NAME: ATKINSON

DATE/TIME OF TEST: 11/14/2017 02:45:51

Diagnostics Check	OK	02:43	
Air Blank	0.000	02:43	
Control Test	0.082	02:43	
Air Blank	0.000	02:44	
Subject Sample	# 1	02:44	Subject Test Refused
Breath Volume	0.000	02:44	
Air Blank	0.000	02:44	
Control Test	0.082	02:45	
Air Blank	0.000	02:45	
Diagnostics Check	OK	02:45	

Data Download Date/Time: 11/17/2017 13:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARRIER,A,R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DAVIS,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:26

SUBJECT'S LAST NAME: MCCLAIN

DATE/TIME OF TEST: 11/14/2017 03:52:27

Diagnostics Check	OK	03:46
Air Blank	0.000	03:46
Control Test	0.082	03:46
Air Blank	0.000	03:47
Subject Sample # 1	0.074	03:48
Breath Volume	2.359	03:48
Air Blank	0.000	03:48
Air Blank	0.000	03:50
Subject Sample # 2	0.077	03:50
Breath Volume	2.605	03:50
Air Blank	0.000	03:51
Control Test	0.081	03:51
Air Blank	0.000	03:52
Diagnostics Check	OK	03:52

Data Download Date/Time: 11/17/2017 13:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DASHER, JOSHUA, L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BROWN, T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 18:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:48

SUBJECT'S LAST NAME: CROSBY

DATE/TIME OF TEST: 11/20/2017 20:15:23

Diagnostics Check	OK	20:09
Air Blank	0.000	20:09
Control Test	0.081	20:09
Air Blank	0.000	20:10
Subject Sample # 1	0.192	20:10
Breath Volume	3.058	20:10
Air Blank	0.000	20:11
Air Blank	0.000	20:13
Subject Sample # 2	0.195	20:13
Breath Volume	2.765	20:13
Air Blank	0.000	20:14
Control Test	0.080	20:14
Air Blank	0.000	20:15
Diagnostics Check	OK	20:15

Data Download Date/Time: 12/11/2017 15:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MACK,CHAVIST,S
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HOOVER,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: HENNER

DATE/TIME OF TEST: 11/21/2017 21:28:54

Diagnostics Check	OK	21:22
Air Blank	0.000	21:22
Control Test	0.081	21:22
Air Blank	0.000	21:23
Subject Sample # 1	0.064	21:23
Breath Volume	1.843	21:23
Air Blank	0.000	21:24
Air Blank	0.000	21:26
Subject Sample # 2	0.062	21:27
Breath Volume	1.941	21:27
Air Blank	0.000	21:27
Control Test	0.081	21:28
Air Blank	0.000	21:28
Diagnostics Check	OK	21:28

Data Download Date/Time: 12/11/2017 15:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY, SHANON, E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HILLIER, J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:37

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 11/24/2017 02:03:34

Diagnostics Check	OK	02:00	
Air Blank	0.000	02:01	
Control Test	0.081	02:01	
Air Blank	0.000	02:02	
Subject Sample	# 1	02:02	Subject Test Refused
Breath Volume	0.000	02:02	
Air Blank	0.000	02:02	
Control Test	0.081	02:02	
Air Blank	0.000	02:03	
Diagnostics Check	OK	02:03	

Data Download Date/Time: 12/11/2017 15:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARRIER,A,R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: STANLEY,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:09

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 11/25/2017 01:43:57

Diagnostics Check	OK	01:30	
Air Blank	0.000	01:30	
Control Test	0.080	01:31	
Air Blank	0.000	01:31	
Subject Sample	# 1 0.119	01:34	Volume Not Met
Breath Volume	0.812	01:34	
Air Blank	0.000	01:35	
Air Blank	0.000	01:37	
Subject Sample	# 2 0.120	01:39	
Breath Volume	1.582	01:39	
Air Blank	0.000	01:39	
Air Blank	0.000	01:41	
Subject Sample	# 3 0.123	01:42	
Breath Volume	1.367	01:42	
Air Blank	0.000	01:43	
Control Test	0.080	01:43	
Air Blank	0.000	01:43	
Diagnostics Check	OK	01:43	

Data Download Date/Time: 12/11/2017 15:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DASHER, JOSHUA, L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: KELLY, C
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 06:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:25

SUBJECT'S LAST NAME: BOATRIGHT

DATE/TIME OF TEST: 11/25/2017 07:48:22

Diagnostics Check	OK	07:45	
Air Blank	0.000	07:45	
Control Test	0.081	07:46	
Air Blank	0.000	07:46	
Subject Sample	# 1	07:47	Subject Test Refused
Breath Volume	0.000	07:47	
Air Blank	0.000	07:47	
Control Test	0.081	07:47	
Air Blank	0.000	07:48	
Diagnostics Check	OK	07:48	

Data Download Date/Time: 12/11/2017 15:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARRIER,A,R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WALLACE,B
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: ANDREW

DATE/TIME OF TEST: 11/26/2017 01:07:36

Diagnostics Check	OK	01:01
Air Blank	0.000	01:01
Control Test	0.081	01:01
Air Blank	0.000	01:02
Subject Sample # 1	0.145	01:03
Breath Volume	1.968	01:03
Air Blank	0.000	01:03
Air Blank	0.000	01:05
Subject Sample # 2	0.136	01:06
Breath Volume	2.351	01:06
Air Blank	0.000	01:06
Control Test	0.081	01:07
Air Blank	0.000	01:07
Diagnostics Check	OK	01:07

Data Download Date/Time: 12/11/2017 15:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARRIER,A,R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BRYANT,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:38

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 11/26/2017 03:05:07

Diagnostics Check	OK	02:58
Air Blank	0.000	02:58
Control Test	0.081	02:59
Air Blank	0.000	02:59
Subject Sample # 1	0.211	03:00
Breath Volume	2.222	03:00
Air Blank	0.000	03:01
Air Blank	0.000	03:02
Subject Sample # 2	0.203	03:03
Breath Volume	3.589	03:03
Air Blank	0.000	03:04
Control Test	0.079	03:04
Air Blank	0.000	03:05
Diagnostics Check	OK	03:05

Data Download Date/Time: 12/11/2017 15:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MACK,C,S
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MOELLER,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:27

SUBJECT'S LAST NAME: EHRHART

DATE/TIME OF TEST: 11/30/2017 19:55:01

Diagnostics Check	OK	19:51	
Air Blank	0.000	19:51	
Control Test	0.081	19:52	
Air Blank	0.000	19:52	
Subject Sample	# 1	19:53	Subject Test Refused
Breath Volume	0.000	19:53	
Air Blank	0.000	19:54	
Control Test	0.081	19:54	
Air Blank	0.000	19:54	
Diagnostics Check	OK	19:55	

Data Download Date/Time: 12/11/2017 15:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001071

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRAUNLICH,G,A
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: FENNICK,T
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 10/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:31

SUBJECT'S LAST NAME: HRINIK

DATE/TIME OF TEST: 11/03/2017 23:59:04

Diagnostics Check	OK	23:52
Air Blank	0.000	23:52
Control Test	0.079	23:52
Air Blank	0.000	23:53
Subject Sample # 1	0.158	23:54
Breath Volume	1.824	23:54
Air Blank	0.000	23:54
Air Blank	0.000	23:56
Subject Sample # 2	0.163	23:57
Breath Volume	2.093	23:57
Air Blank	0.000	23:58
Control Test	0.079	23:58
Air Blank	0.000	23:58
Diagnostics Check	OK	23:59

Data Download Date/Time: 11/21/2017 11:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001071

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRAUNLICH,G,A
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: MILLS,D
ARRESTING OFFICER AGENCY: Orange Park PD
LAST AGENCY INSPECTION: 10/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:31

SUBJECT'S LAST NAME: LHOMMEDIEU

DATE/TIME OF TEST: 11/05/2017 01:58:05

Diagnostics Check	OK	01:51
Air Blank	0.000	01:51
Control Test	0.079	01:51
Air Blank	0.000	01:52
Subject Sample # 1	0.144	01:53
Breath Volume	3.308	01:53
Air Blank	0.000	01:53
Air Blank	0.000	01:55
Subject Sample # 2	0.142	01:56
Breath Volume	1.394	01:56
Air Blank	0.000	01:57
Control Test	0.080	01:57
Air Blank	0.000	01:58
Diagnostics Check	OK	01:58

Data Download Date/Time: 11/21/2017 11:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001071

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL, DAKOTA, W
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: LANIER, H
ARRESTING OFFICER AGENCY: Orange Park PD
LAST AGENCY INSPECTION: 10/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:51

SUBJECT'S LAST NAME: BURCHETT

DATE/TIME OF TEST: 11/09/2017 05:18:15

Diagnostics Check	OK	05:12
Air Blank	0.000	05:12
Control Test	0.079	05:12
Air Blank	0.000	05:13
Subject Sample # 1	0.155	05:13
Breath Volume	1.820	05:13
Air Blank	0.000	05:14
Air Blank	0.000	05:16
Subject Sample # 2	0.159	05:16
Breath Volume	1.796	05:16
Air Blank	0.000	05:17
Control Test	0.080	05:17
Air Blank	0.000	05:18
Diagnostics Check	OK	05:18

Data Download Date/Time: 11/21/2017 11:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001071

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALTSTATT,MATTHEW,B
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: BRIGGS,G
ARRESTING OFFICER AGENCY: Orange Park PD
LAST AGENCY INSPECTION: 10/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 18:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:05

SUBJECT'S LAST NAME: FARHAT

DATE/TIME OF TEST: 11/12/2017 20:37:50

Diagnostics Check	OK	20:25	
Air Blank	0.000	20:25	
Control Test	0.080	20:26	
Air Blank	0.000	20:26	
Subject Sample	# 1	20:29	No Sample Provided
Breath Volume	0.000	20:29	
Air Blank	0.000	20:30	
Air Blank	0.000	20:32	
Subject Sample	# 2	0.111	20:33
Breath Volume	1.468	20:33	
Air Blank	0.000	20:33	
Air Blank	0.000	20:35	
Subject Sample	# 3	0.112	20:36
Breath Volume	1.613	20:36	
Air Blank	0.000	20:36	
Control Test	0.080	20:37	
Air Blank	0.000	20:37	
Diagnostics Check	OK	20:37	

Data Download Date/Time: 11/21/2017 11:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001071

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRAUNLICH,G,A
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: ALEXANDER,D
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 10/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:32

SUBJECT'S LAST NAME: BOYER

DATE/TIME OF TEST: 11/18/2017 03:01:36

Diagnostics Check	OK	02:54
Air Blank	0.000	02:54
Control Test	0.079	02:55
Air Blank	0.000	02:55
Subject Sample # 1	0.186	02:57
Breath Volume	1.949	02:57
Air Blank	0.000	02:57
Air Blank	0.000	02:59
Subject Sample # 2	0.171	03:00
Breath Volume	1.187	03:00
Air Blank	0.000	03:00
Control Test	0.079	03:01
Air Blank	0.000	03:01
Diagnostics Check	OK	03:01

Data Download Date/Time: 11/21/2017 11:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001071

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL, DAKOTA, W
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: COLLINS, K
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 10/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:34

SUBJECT'S LAST NAME: IMLER

DATE/TIME OF TEST: 11/19/2017 02:01:03

Diagnostics Check	OK	01:54
Air Blank	0.000	01:54
Control Test	0.080	01:55
Air Blank	0.000	01:55
Subject Sample # 1	0.168	01:56
Breath Volume	2.652	01:56
Air Blank	0.000	01:57
Air Blank	0.000	01:59
Subject Sample # 2	0.168	01:59
Breath Volume	2.441	01:59
Air Blank	0.000	02:00
Control Test	0.079	02:00
Air Blank	0.000	02:00
Diagnostics Check	OK	02:01

Data Download Date/Time: 11/21/2017 11:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001071

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANTORO,CAMERON,J
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: BRED,CECIL
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 11/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:15

SUBJECT'S LAST NAME: BRED

DATE/TIME OF TEST: 11/30/2017 04:45:22

Diagnostics Check	OK	04:38
Air Blank	0.000	04:39
Control Test	0.080	04:39
Air Blank	0.000	04:40
Subject Sample # 1	0.238	04:40
Breath Volume	2.945	04:40
Air Blank	0.000	04:41
Air Blank	0.000	04:43
Subject Sample # 2	0.242	04:43
Breath Volume	3.066	04:43
Air Blank	0.000	04:44
Control Test	0.080	04:44
Air Blank	0.000	04:45
Diagnostics Check	OK	04:45

Data Download Date/Time: 12/23/2017 16:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001071

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANTORO,CAMERON,J
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: TENNANT,R
ARRESTING OFFICER AGENCY: Orange Park PD
LAST AGENCY INSPECTION: 11/21/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:25

SUBJECT'S LAST NAME: ORTEGA

DATE/TIME OF TEST: 11/30/2017 05:54:48

Diagnostics Check	OK	05:46
Air Blank	0.000	05:47
Control Test	0.079	05:47
Air Blank	0.000	05:48
Subject Sample # 1	0.189	05:48
Breath Volume	2.011	05:48
Air Blank	0.000	05:49
Air Blank	0.000	05:51
Subject Sample # 2	0.175	05:53
Breath Volume	1.832	05:53
Air Blank	0.000	05:53
Control Test	0.080	05:54
Air Blank	0.000	05:54
Diagnostics Check	OK	05:54

Data Download Date/Time: 12/23/2017 16:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001071

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANTORO,CAMERON,J
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: COX,Z
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 11/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 19:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:30

SUBJECT'S LAST NAME: MEECE

DATE/TIME OF TEST: 11/30/2017 20:58:17

Diagnostics Check	OK	20:51
Air Blank	0.000	20:52
Control Test	0.079	20:52
Air Blank	0.000	20:53
Subject Sample # 1	0.179	20:53
Breath Volume	4.710	20:53
Air Blank	0.000	20:54
Air Blank	0.000	20:56
Subject Sample # 2	0.172	20:56
Breath Volume	1.765	20:56
Air Blank	0.000	20:57
Control Test	0.079	20:57
Air Blank	0.000	20:58
Diagnostics Check	OK	20:58

Data Download Date/Time: 12/23/2017 16:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001074

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,JASON,G
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: BROWN,JASON
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 10/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: CUMBIE

DATE/TIME OF TEST: 11/11/2017 01:35:40

Diagnostics Check	OK	01:29
Air Blank	0.000	01:29
Control Test	0.079	01:30
Air Blank	0.000	01:30
Subject Sample # 1	0.121	01:31
Breath Volume	1.738	01:31
Air Blank	0.000	01:31
Air Blank	0.000	01:33
Subject Sample # 2	0.123	01:34
Breath Volume	2.843	01:34
Air Blank	0.000	01:34
Control Test	0.078	01:35
Air Blank	0.000	01:35
Diagnostics Check	OK	01:35

Data Download Date/Time: 11/29/2017 15:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001074

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAUS,SCOTT,G
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: MAUS,SCOTT
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 10/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 15:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:02

SUBJECT'S LAST NAME: ARROYO

DATE/TIME OF TEST: 11/11/2017 17:28:54

Diagnostics Check	OK	17:22
Air Blank	0.000	17:23
Control Test	0.079	17:23
Air Blank	0.000	17:24
Subject Sample # 1	0.000	17:24
Breath Volume	4.285	17:24
Air Blank	0.000	17:25
Air Blank	0.000	17:27
Subject Sample # 2	0.000	17:27
Breath Volume	3.640	17:27
Air Blank	0.000	17:27
Control Test	0.079	17:28
Air Blank	0.000	17:28
Diagnostics Check	OK	17:28

Data Download Date/Time: 11/29/2017 15:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001074

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARD,STEVEN,J
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: WARD,STEVEN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 10/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: GORDON

DATE/TIME OF TEST: 11/18/2017 01:27:52

Diagnostics Check	OK	01:21
Air Blank	0.000	01:21
Control Test	0.079	01:21
Air Blank	0.000	01:22
Subject Sample # 1	0.190	01:23
Breath Volume	3.863	01:23
Air Blank	0.000	01:23
Air Blank	0.000	01:25
Subject Sample # 2	0.192	01:26
Breath Volume	3.437	01:26
Air Blank	0.000	01:26
Control Test	0.077	01:27
Air Blank	0.000	01:27
Diagnostics Check	OK	01:27

Data Download Date/Time: 11/29/2017 15:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001074

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COFFEE,JASON,A
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: COFFEE,JASON
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 10/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 11/19/2017 01:48:44

Diagnostics Check	OK	01:37	
Air Blank	0.000	01:37	
Control Test	0.079	01:38	
Air Blank	0.000	01:38	
Subject Sample	# 1	01:41	No Sample Provided
Breath Volume	0.000	01:41	
Air Blank	0.000	01:42	
Air Blank	0.000	01:44	
Subject Sample	# 2	01:47	No Sample Provided
Breath Volume	0.000	01:47	
Air Blank	0.000	01:47	
Control Test	0.078	01:48	
Air Blank	0.000	01:48	
Diagnostics Check	OK	01:48	

Data Download Date/Time: 11/29/2017 15:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001074

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COFFEE,JASON,A
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: COFFEE,JASON
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 10/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: BERGERON

DATE/TIME OF TEST: 11/23/2017 00:18:08

Diagnostics Check	OK	00:11
Air Blank	0.000	00:11
Control Test	0.078	00:12
Air Blank	0.000	00:12
Subject Sample # 1	0.190	00:13
Breath Volume	5.128	00:13
Air Blank	0.000	00:14
Air Blank	0.000	00:15
Subject Sample # 2	0.183	00:16
Breath Volume	4.191	00:16
Air Blank	0.000	00:17
Control Test	0.077	00:17
Air Blank	0.000	00:18
Diagnostics Check	OK	00:18

Data Download Date/Time: 11/29/2017 15:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001074

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COFFEE,JASON,A
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: COFFEE,JASON
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 10/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 19:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: TOWNSEND

DATE/TIME OF TEST: 11/23/2017 21:22:54

Diagnostics Check	OK	21:11	
Air Blank	0.000	21:11	
Control Test	0.078	21:12	
Air Blank	0.000	21:12	
Subject Sample	# 1	21:15	No Sample Provided
Breath Volume	0.000	21:15	
Air Blank	0.000	21:16	
Air Blank	0.000	21:18	
Subject Sample	# 2	21:21	No Sample Provided
Breath Volume	0.000	21:21	
Air Blank	0.000	21:21	
Control Test	0.078	21:22	
Air Blank	0.000	21:22	
Diagnostics Check	OK	21:22	

Data Download Date/Time: 11/29/2017 15:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001074

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,JASON,G
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: BROWN,JASON
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 10/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: HENRY

DATE/TIME OF TEST: 11/26/2017 01:21:23

Diagnostics Check	OK	01:14
Air Blank	0.000	01:15
Control Test	0.079	01:15
Air Blank	0.000	01:16
Subject Sample # 1	0.227	01:16
Breath Volume	1.433	01:16
Air Blank	0.000	01:17
Air Blank	0.000	01:19
Subject Sample # 2	0.233	01:19
Breath Volume	1.714	01:19
Air Blank	0.000	01:20
Control Test	0.078	01:20
Air Blank	0.000	01:21
Diagnostics Check	OK	01:21

Data Download Date/Time: 11/29/2017 15:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001074

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRIOLAT,CASEY,J
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: BRIOLAT,CASEY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 10/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:42

SUBJECT'S LAST NAME: GARCIA CABRERA

DATE/TIME OF TEST: 11/28/2017 05:11:40

Diagnostics Check	OK	05:03
Air Blank	0.000	05:03
Control Test	0.078	05:04
Air Blank	0.000	05:04
Subject Sample # 1	0.210	05:06
Breath Volume	1.210	05:06
Air Blank	0.000	05:06
Air Blank	0.000	05:08
Subject Sample # 2	0.217	05:09
Breath Volume	1.468	05:09
Air Blank	0.000	05:10
Control Test	0.077	05:11
Air Blank	0.000	05:11
Diagnostics Check	OK	05:11

Data Download Date/Time: 11/29/2017 15:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001076

INSTRUMENT REGISTERED TO: ZEPHYRHILLS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HILLEN,MATTHEW,J
BREATH TEST OPERATOR AGENCY: Zephyrhills PD
ARRESTING OFFICER: HILLEN,MATTHEW
ARRESTING OFFICER AGENCY: Zephyrhills PD
LAST AGENCY INSPECTION: 10/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 11:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:17

SUBJECT'S LAST NAME: WARD

DATE/TIME OF TEST: 11/16/2017 11:44:34

Diagnostics Check	OK	11:38
Air Blank	0.000	11:39
Control Test	0.078	11:39
Air Blank	0.000	11:40
Subject Sample # 1	0.000	11:40
Breath Volume	4.085	11:40
Air Blank	0.000	11:40
Air Blank	0.000	11:42
Subject Sample # 2	0.000	11:43
Breath Volume	3.367	11:43
Air Blank	0.000	11:43
Control Test	0.078	11:43
Air Blank	0.000	11:44
Diagnostics Check	OK	11:44

Data Download Date/Time: 11/28/2017 08:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZOLLOTA,VLADIMIR,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ALLI,ANTHONY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 11/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: BRYANT

DATE/TIME OF TEST: 11/03/2017 00:22:52

Diagnostics Check	OK	00:20	
Air Blank	0.000	00:20	
Control Test	0.080	00:20	
Air Blank	0.000	00:21	
Subject Sample	# 1	00:21	Subject Test Refused
Breath Volume	0.000	00:21	
Air Blank	0.000	00:21	
Control Test	0.079	00:22	
Air Blank	0.000	00:22	
Diagnostics Check	OK	00:22	

Data Download Date/Time: 12/06/2017 10:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZOLLOTA,VLADIMIR,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: KARAYIANES,MICHAEL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 11/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:16

SUBJECT'S LAST NAME: POTTS

DATE/TIME OF TEST: 11/04/2017 01:48:24

Diagnostics Check	OK	01:42
Air Blank	0.000	01:42
Control Test	0.080	01:42
Air Blank	0.000	01:43
Subject Sample # 1	0.076	01:43
Breath Volume	2.984	01:43
Air Blank	0.000	01:44
Air Blank	0.000	01:46
Subject Sample # 2	0.074	01:46
Breath Volume	3.117	01:46
Air Blank	0.000	01:47
Control Test	0.079	01:47
Air Blank	0.000	01:48
Diagnostics Check	OK	01:48

Data Download Date/Time: 12/06/2017 10:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZOLLOTA,VLADIMIR,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ELLIS,ALEX
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: CARDOSO

DATE/TIME OF TEST: 11/04/2017 02:10:52

Diagnostics Check	OK	02:04
Air Blank	0.000	02:04
Control Test	0.080	02:05
Air Blank	0.000	02:05
Subject Sample # 1	0.107	02:06
Breath Volume	2.539	02:06
Air Blank	0.000	02:06
Air Blank	0.000	02:08
Subject Sample # 2	0.102	02:09
Breath Volume	1.753	02:09
Air Blank	0.000	02:09
Control Test	0.079	02:10
Air Blank	0.000	02:10
Diagnostics Check	OK	02:10

Data Download Date/Time: 12/06/2017 10:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZOLLOTA,VLADIMIR,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: KARAYIANES,MICHAEL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 11/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: FARLEY

DATE/TIME OF TEST: 11/04/2017 22:23:25

Diagnostics Check	OK	22:16
Air Blank	0.000	22:17
Control Test	0.079	22:17
Air Blank	0.000	22:18
Subject Sample # 1	0.368	22:18
Breath Volume	1.261	22:18
Air Blank	0.000	22:19
Air Blank	0.000	22:21
Subject Sample # 2	0.359	22:21
Breath Volume	1.738	22:21
Air Blank	0.000	22:22
Control Test	0.078	22:22
Air Blank	0.000	22:23
Diagnostics Check	OK	22:23

Data Download Date/Time: 12/06/2017 10:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZOLLOTA,VLADIMIR,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: COX,LENARD
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 11/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 11/05/2017 01:58:01

Diagnostics Check	OK	01:51
Air Blank	0.000	01:51
Control Test	0.080	01:51
Air Blank	0.000	01:52
Subject Sample # 1	0.293	01:52
Breath Volume	2.011	01:52
Air Blank	0.000	01:53
Air Blank	0.000	01:55
Subject Sample # 2	0.280	01:56
Breath Volume	1.253	01:56
Air Blank	0.000	01:57
Control Test	0.079	01:57
Air Blank	0.000	01:57
Diagnostics Check	OK	01:58

Data Download Date/Time: 12/06/2017 10:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZOLLOTA,VLADIMIR,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ALLI,ANTHONY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 11/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: CALDERON

DATE/TIME OF TEST: 11/05/2017 03:08:23

Diagnostics Check	OK	03:05	
Air Blank	0.000	03:05	
Control Test	0.079	03:06	
Air Blank	0.000	03:06	
Subject Sample	# 1	03:07	Subject Test Refused
Breath Volume	0.000	03:07	
Air Blank	0.000	03:07	
Control Test	0.079	03:07	
Air Blank	0.000	03:08	
Diagnostics Check	OK	03:08	

Data Download Date/Time: 12/06/2017 10:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLI, ANTHONY,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ZOLLOTA, VLADIMIR
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 11/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: SCHAAF

DATE/TIME OF TEST: 11/08/2017 23:12:25

Diagnostics Check	OK	22:58	
Air Blank	0.000	22:59	
Control Test	0.079	22:59	
Air Blank	0.000	23:00	
Subject Sample	# 1	23:03	No Sample Provided
Breath Volume	0.000	23:03	
Air Blank	0.000	23:03	
Air Blank	0.000	23:05	
Subject Sample	# 2	0.140	23:07
Breath Volume	2.769	23:07	
Air Blank	0.000	23:08	
Air Blank	0.000	23:10	
Subject Sample	# 3	0.141	23:10
Breath Volume	2.117	23:10	
Air Blank	0.000	23:11	
Control Test	0.079	23:11	
Air Blank	0.000	23:12	
Diagnostics Check	OK	23:12	

Data Download Date/Time: 12/06/2017 10:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZOLLOTA,VLADIMIR,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: KARAYIANES,MICHAEL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 11/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:49

SUBJECT'S LAST NAME: HOTH

DATE/TIME OF TEST: 11/10/2017 23:34:32

Diagnostics Check	OK	23:28
Air Blank	0.000	23:29
Control Test	0.080	23:29
Air Blank	0.000	23:30
Subject Sample # 1	0.000	23:30
Breath Volume	2.449	23:30
Air Blank	0.000	23:30
Air Blank	0.000	23:32
Subject Sample # 2	0.000	23:33
Breath Volume	2.144	23:33
Air Blank	0.000	23:33
Control Test	0.079	23:33
Air Blank	0.000	23:34
Diagnostics Check	OK	23:34

Data Download Date/Time: 12/06/2017 10:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZOLLOTA,VLADIMIR,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: WALDO,AARON
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 11/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:48

SUBJECT'S LAST NAME: HARDEE

DATE/TIME OF TEST: 11/12/2017 00:20:36

Diagnostics Check	OK	00:14
Air Blank	0.000	00:14
Control Test	0.080	00:15
Air Blank	0.000	00:15
Subject Sample # 1	0.150	00:16
Breath Volume	2.308	00:16
Air Blank	0.000	00:16
Air Blank	0.000	00:18
Subject Sample # 2	0.145	00:19
Breath Volume	2.488	00:19
Air Blank	0.000	00:19
Control Test	0.079	00:20
Air Blank	0.000	00:20
Diagnostics Check	OK	00:20

Data Download Date/Time: 12/06/2017 10:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KARAYIANES,MICHAEL,F
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ZOLLOTA,VLADIMIR
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 11/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: BUENO

DATE/TIME OF TEST: 11/12/2017 01:37:56

Diagnostics Check	OK	01:31
Air Blank	0.000	01:32
Control Test	0.079	01:32
Air Blank	0.000	01:32
Subject Sample # 1	0.185	01:33
Breath Volume	1.683	01:33
Air Blank	0.000	01:34
Air Blank	0.000	01:35
Subject Sample # 2	0.177	01:36
Breath Volume	2.195	01:36
Air Blank	0.000	01:37
Control Test	0.078	01:37
Air Blank	0.000	01:37
Diagnostics Check	OK	01:37

Data Download Date/Time: 12/06/2017 10:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLI, ANTHONY,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: KARAYIANES, MICHAEL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 11/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:13

SUBJECT'S LAST NAME: MCGRUFF

DATE/TIME OF TEST: 11/17/2017 03:06:12

Diagnostics Check	OK	02:59
Air Blank	0.000	02:59
Control Test	0.080	03:00
Air Blank	0.000	03:00
Subject Sample # 1	0.071	03:01
Breath Volume	3.347	03:01
Air Blank	0.000	03:02
Air Blank	0.000	03:04
Subject Sample # 2	0.071	03:04
Breath Volume	3.820	03:04
Air Blank	0.000	03:05
Control Test	0.080	03:05
Air Blank	0.000	03:06
Diagnostics Check	OK	03:06

Data Download Date/Time: 12/06/2017 10:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZOLLOTA,VLADIMIR,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: KARAYIANES,MICHAEL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 11/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:22

SUBJECT'S LAST NAME: SKAGGS

DATE/TIME OF TEST: 11/18/2017 03:08:10

Diagnostics Check	OK	03:05	
Air Blank	0.000	03:05	
Control Test	0.080	03:06	
Air Blank	0.000	03:06	
Subject Sample	# 1	03:06	Subject Test Refused
Breath Volume	0.000	03:06	
Air Blank	0.000	03:07	
Control Test	0.080	03:07	
Air Blank	0.000	03:08	
Diagnostics Check	OK	03:08	

Data Download Date/Time: 12/06/2017 10:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KARAYIANES,MICHAEL,F
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ZOLLOTA,VLADIMIR
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 11/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:50

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 11/23/2017 19:35:49

Diagnostics Check	OK	19:29
Air Blank	0.000	19:29
Control Test	0.079	19:29
Air Blank	0.000	19:30
Subject Sample # 1	0.335	19:30
Breath Volume	3.582	19:30
Air Blank	0.000	19:31
Air Blank	0.000	19:33
Subject Sample # 2	0.342	19:34
Breath Volume	3.652	19:34
Air Blank	0.000	19:34
Control Test	0.078	19:35
Air Blank	0.000	19:35
Diagnostics Check	OK	19:35

Data Download Date/Time: 12/06/2017 10:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: SILVERSTEIN,CHRISTINA

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 10/05/2017

Date or Time Changed: No

DRY GAS STANDARD LOT #: 32316080A3

DRY GAS EXPIRY DATE: 01/05/2019

VIOLATION CODE: DUI

ARREST TIME: 02:17

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:36

SUBJECT'S LAST NAME: HARVEY

DATE/TIME OF TEST: 11/03/2017 03:02:45

Diagnostics Check	OK	02:56
Air Blank	0.000	02:56
Control Test	0.079	02:56
Air Blank	0.000	02:57
Subject Sample # 1	0.192	02:58
Breath Volume	2.390	02:58
Air Blank	0.000	02:58
Air Blank	0.000	03:00
Subject Sample # 2	0.193	03:01
Breath Volume	2.402	03:01
Air Blank	0.000	03:01
Control Test	0.080	03:02
Air Blank	0.000	03:02
Diagnostics Check	OK	03:02

Data Download Date/Time: 11/22/2017 15:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HASTINGS,KENNETH,S

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: HASTINGS,KENNETH

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 10/05/2017

Date or Time Changed: No

DRY GAS STANDARD LOT #: 32316080A3

DRY GAS EXPIRY DATE: 01/05/2019

VIOLATION CODE: DUI

ARREST TIME: 00:36

GENDER: Female

OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: WALKER

DATE/TIME OF TEST: 11/05/2017 01:28:24

Diagnostics Check	OK	01:21
Air Blank	0.000	01:21
Control Test	0.080	01:21
Air Blank	0.000	01:22
Subject Sample # 1	0.232	01:23
Breath Volume	2.203	01:23
Air Blank	0.000	01:23
Air Blank	0.000	01:25
Subject Sample # 2	0.217	01:26
Breath Volume	2.582	01:26
Air Blank	0.000	01:27
Control Test	0.080	01:27
Air Blank	0.000	01:28
Diagnostics Check	OK	01:28

Data Download Date/Time: 11/22/2017 15:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: SILVERSTEIN,CHRISTINA

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 10/05/2017

Date or Time Changed: No

DRY GAS STANDARD LOT #: 32316080A3

DRY GAS EXPIRY DATE: 01/05/2019

VIOLATION CODE: DUI

ARREST TIME: 03:27

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:43

SUBJECT'S LAST NAME: LODER

DATE/TIME OF TEST: 11/05/2017 04:08:22

Diagnostics Check	OK	04:05	
Air Blank	0.000	04:05	
Control Test	0.081	04:06	
Air Blank	0.000	04:06	
Subject Sample	# 1	04:07	Subject Test Refused
Breath Volume	0.000	04:07	
Air Blank	0.000	04:07	
Control Test	0.080	04:07	
Air Blank	0.000	04:08	
Diagnostics Check	OK	04:08	

Data Download Date/Time: 11/22/2017 15:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: SILVERSTEIN,CHRISTINA

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 10/05/2017

Date or Time Changed: No

DRY GAS STANDARD LOT #: 32316080A3

DRY GAS EXPIRY DATE: 01/05/2019

VIOLATION CODE: DUI

ARREST TIME: 18:55

GENDER: Male

OBSERVATION PERIOD BEGAN: 19:28

SUBJECT'S LAST NAME: DOERBECKER

DATE/TIME OF TEST: 11/09/2017 19:54:58

Diagnostics Check	OK	19:48
Air Blank	0.000	19:49
Control Test	0.080	19:49
Air Blank	0.000	19:50
Subject Sample # 1	0.194	19:50
Breath Volume	2.675	19:50
Air Blank	0.000	19:51
Air Blank	0.000	19:52
Subject Sample # 2	0.192	19:53
Breath Volume	2.210	19:53
Air Blank	0.000	19:54
Control Test	0.079	19:54
Air Blank	0.000	19:54
Diagnostics Check	OK	19:54

Data Download Date/Time: 11/22/2017 15:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: SILVERSTEIN,CHRISTINA

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 10/05/2017

Date or Time Changed: No

DRY GAS STANDARD LOT #: 32316080A3

DRY GAS EXPIRY DATE: 01/05/2019

VIOLATION CODE: DUI

ARREST TIME: 22:42

GENDER: Male

OBSERVATION PERIOD BEGAN: 22:58

SUBJECT'S LAST NAME: GARCIA-SEBASTIAN

DATE/TIME OF TEST: 11/09/2017 23:25:44

Diagnostics Check	OK	23:19
Air Blank	0.000	23:19
Control Test	0.079	23:19
Air Blank	0.000	23:20
Subject Sample # 1	0.159	23:21
Breath Volume	1.304	23:21
Air Blank	0.000	23:21
Air Blank	0.000	23:23
Subject Sample # 2	0.161	23:24
Breath Volume	1.707	23:24
Air Blank	0.000	23:24
Control Test	0.079	23:25
Air Blank	0.000	23:25
Diagnostics Check	OK	23:25

Data Download Date/Time: 11/22/2017 15:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUNT,JERRY,D
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: GARCIA,MARVIN
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 10/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 15:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:13

SUBJECT'S LAST NAME: LYONS

DATE/TIME OF TEST: 11/15/2017 16:43:24

Diagnostics Check	OK	16:37
Air Blank	0.000	16:37
Control Test	0.080	16:37
Air Blank	0.000	16:38
Subject Sample # 1	0.000	16:39
Breath Volume	2.406	16:39
Air Blank	0.000	16:39
Air Blank	0.000	16:41
Subject Sample # 2	0.000	16:42
Breath Volume	2.097	16:42
Air Blank	0.000	16:42
Control Test	0.080	16:42
Air Blank	0.000	16:43
Diagnostics Check	OK	16:43

Data Download Date/Time: 11/22/2017 15:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HASTINGS,KENNETH,S

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: 4L26,HASTINGS

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 10/05/2017

Date or Time Changed: No

DRY GAS STANDARD LOT #: 32316080A3

DRY GAS EXPIRY DATE: 01/05/2019

VIOLATION CODE: DUI

ARREST TIME: 20:17

GENDER: Male

OBSERVATION PERIOD BEGAN: 20:33

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 11/15/2017 20:58:53

Diagnostics Check	OK	20:53	
Air Blank	0.000	20:54	
Control Test	0.080	20:54	
Air Blank	0.000	20:55	
Subject Sample	# 1	20:57	Subject Test Refused
Breath Volume	0.000	20:57	
Air Blank	0.000	20:57	
Control Test	0.080	20:58	
Air Blank	0.000	20:58	
Diagnostics Check	OK	20:58	

Data Download Date/Time: 11/22/2017 15:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: SILVERSTEIN,CHRISTINA

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 10/05/2017

Date or Time Changed: No

DRY GAS STANDARD LOT #: 32316080A3

DRY GAS EXPIRY DATE: 01/05/2019

VIOLATION CODE: DUI

ARREST TIME: 02:42

GENDER: Female

OBSERVATION PERIOD BEGAN: 02:57

SUBJECT'S LAST NAME: MORRIS

DATE/TIME OF TEST: 11/16/2017 03:20:42

Diagnostics Check	OK	03:17	
Air Blank	0.000	03:18	
Control Test	0.080	03:18	
Air Blank	0.000	03:19	
Subject Sample	# 1	03:19	Subject Test Refused
Breath Volume	0.000	03:19	
Air Blank	0.000	03:19	
Control Test	0.080	03:20	
Air Blank	0.000	03:20	
Diagnostics Check	OK	03:20	

Data Download Date/Time: 11/22/2017 15:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: THORNSBROUGH,DREW,A

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: THORNSBROUGH,DREW

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 10/05/2017

Date or Time Changed: No

DRY GAS STANDARD LOT #: 32316080A3

DRY GAS EXPIRY DATE: 01/05/2019

VIOLATION CODE: DUI

ARREST TIME: 08:51

GENDER: Female

OBSERVATION PERIOD BEGAN: 09:00

SUBJECT'S LAST NAME: SAVOIE

DATE/TIME OF TEST: 11/16/2017 09:24:11

Diagnostics Check	OK	09:20	
Air Blank	0.000	09:20	
Control Test	0.081	09:21	
Air Blank	0.000	09:21	
Subject Sample	# 1	09:22	Subject Test Refused
Breath Volume	0.000	09:22	
Air Blank	0.000	09:23	
Control Test	0.080	09:23	
Air Blank	0.000	09:24	
Diagnostics Check	OK	09:24	

Data Download Date/Time: 11/22/2017 15:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALAMEDA,JORGE,L
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: ALAMEDA,JORGE
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 10/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 18:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:35

SUBJECT'S LAST NAME: STANLEY

DATE/TIME OF TEST: 11/16/2017 19:08:16

Diagnostics Check	OK	19:01
Air Blank	0.000	19:02
Control Test	0.080	19:02
Air Blank	0.000	19:02
Subject Sample # 1	0.173	19:03
Breath Volume	3.828	19:03
Air Blank	0.000	19:04
Air Blank	0.000	19:06
Subject Sample # 2	0.170	19:06
Breath Volume	1.960	19:06
Air Blank	0.000	19:07
Control Test	0.080	19:07
Air Blank	0.000	19:08
Diagnostics Check	OK	19:08

Data Download Date/Time: 11/22/2017 15:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUNT,JERRY,D
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: BREWER,CHRISTINA
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 10/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:53

SUBJECT'S LAST NAME: LASCH

DATE/TIME OF TEST: 11/17/2017 03:20:25

Diagnostics Check	OK	03:13
Air Blank	0.000	03:14
Control Test	0.081	03:14
Air Blank	0.000	03:15
Subject Sample # 1	0.166	03:15
Breath Volume	1.804	03:15
Air Blank	0.000	03:16
Air Blank	0.000	03:18
Subject Sample # 2	0.165	03:18
Breath Volume	1.558	03:18
Air Blank	0.000	03:19
Control Test	0.080	03:19
Air Blank	0.000	03:20
Diagnostics Check	OK	03:20

Data Download Date/Time: 11/22/2017 15:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUNT,JERRY,D
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: BREWER,CHRISTINA
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 10/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: DELLAVECCHIA

DATE/TIME OF TEST: 11/17/2017 23:28:57

Diagnostics Check	OK	23:22
Air Blank	0.000	23:23
Control Test	0.080	23:23
Air Blank	0.000	23:23
Subject Sample # 1	0.000	23:24
Breath Volume	4.199	23:24
Air Blank	0.000	23:25
Air Blank	0.000	23:27
Subject Sample # 2	0.000	23:27
Breath Volume	3.386	23:27
Air Blank	0.000	23:28
Control Test	0.081	23:28
Air Blank	0.000	23:28
Diagnostics Check	OK	23:28

Data Download Date/Time: 11/22/2017 15:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 10/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: KINCHELOE

DATE/TIME OF TEST: 11/18/2017 02:54:19

Diagnostics Check	OK	02:51	
Air Blank	0.000	02:51	
Control Test	0.081	02:51	
Air Blank	0.000	02:52	
Subject Sample	# 1	02:52	Subject Test Refused
Breath Volume	0.000	02:52	
Air Blank	0.000	02:53	
Control Test	0.080	02:53	
Air Blank	0.000	02:54	
Diagnostics Check	OK	02:54	

Data Download Date/Time: 11/22/2017 15:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 10/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: BUENROSTRO

DATE/TIME OF TEST: 11/19/2017 01:31:55

Diagnostics Check	OK	01:23	
Air Blank	0.000	01:23	
Control Test	0.080	01:24	
Air Blank	0.000	01:24	
Subject Sample	# 1	01:27	No Sample Provided
Breath Volume	0.000	01:27	
Air Blank	0.000	01:28	
Air Blank	0.000	01:30	
Subject Sample	# 2	01:30	Subject Test Refused
Breath Volume	0.000	01:30	
Air Blank	0.000	01:31	
Control Test	0.080	01:31	
Air Blank	0.000	01:31	
Diagnostics Check	OK	01:31	

Data Download Date/Time: 11/22/2017 15:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HASTINGS,KENNETH,S

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: HASTINGS,KENNETH

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 10/05/2017

Date or Time Changed: No

DRY GAS STANDARD LOT #: 32316080A3

DRY GAS EXPIRY DATE: 01/05/2019

VIOLATION CODE: DUI

ARREST TIME: 02:06

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:17

SUBJECT'S LAST NAME: LINDAHL

DATE/TIME OF TEST: 11/19/2017 02:41:44

Diagnostics Check	OK	02:37	
Air Blank	0.000	02:38	
Control Test	0.081	02:38	
Air Blank	0.000	02:38	
Subject Sample	# 1	02:40	Subject Test Refused
Breath Volume	0.000	02:40	
Air Blank	0.000	02:40	
Control Test	0.081	02:41	
Air Blank	0.000	02:41	
Diagnostics Check	OK	02:41	

Data Download Date/Time: 11/22/2017 15:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: WILSON,CHRISTOPHER,R

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: GARCIA,M

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 11/21/2017

Date or Time Changed: No

DRY GAS STANDARD LOT #: 32316080A3

DRY GAS EXPIRY DATE: 01/05/2019

VIOLATION CODE: DUI

ARREST TIME: 11:43

GENDER: Male

OBSERVATION PERIOD BEGAN: 12:13

SUBJECT'S LAST NAME: BRITTON

DATE/TIME OF TEST: 11/22/2017 12:43:15

Diagnostics Check	OK	12:34
Air Blank	0.000	12:35
Control Test	0.081	12:35
Air Blank	0.000	12:36
Subject Sample # 1	0.000	12:38
Breath Volume	1.613	12:38
Air Blank	0.000	12:38
Air Blank	0.000	12:40
Subject Sample # 2	0.000	12:41
Breath Volume	1.796	12:41
Air Blank	0.000	12:42
Control Test	0.080	12:42
Air Blank	0.000	12:43
Diagnostics Check	OK	12:43

Data Download Date/Time: 11/22/2017 15:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: SILVERSTEIN,CHRISTINA

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 11/21/2017

Date or Time Changed: No

DRY GAS STANDARD LOT #: 32316080A3

DRY GAS EXPIRY DATE: 01/05/2019

VIOLATION CODE: DUI

ARREST TIME: 02:06

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:23

SUBJECT'S LAST NAME: DUNSIZER

DATE/TIME OF TEST: 11/25/2017 02:46:33

Diagnostics Check	OK	02:43	
Air Blank	0.000	02:44	
Control Test	0.080	02:44	
Air Blank	0.000	02:44	
Subject Sample	# 1	02:45	Subject Test Refused
Breath Volume	0.000	02:45	
Air Blank	0.000	02:45	
Control Test	0.080	02:45	
Air Blank	0.000	02:46	
Diagnostics Check	OK	02:46	

Data Download Date/Time: 12/18/2017 17:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HASTINGS,KENNETH,S

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: HASTINGS,KENNETH

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 11/21/2017

Date or Time Changed: No

DRY GAS STANDARD LOT #: 32316080A3

DRY GAS EXPIRY DATE: 01/05/2019

VIOLATION CODE: DUI

ARREST TIME: 19:21

GENDER: Male

OBSERVATION PERIOD BEGAN: 19:34

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 11/25/2017 20:01:09

Diagnostics Check	OK	19:54
Air Blank	0.000	19:54
Control Test	0.080	19:55
Air Blank	0.000	19:55
Subject Sample # 1	0.440	19:56
Breath Volume	5.390	19:56
Air Blank	0.000	19:57
Air Blank	0.000	19:58
Subject Sample # 2	0.424	19:59
Breath Volume	5.117	19:59
Air Blank	0.000	20:00
Control Test	0.080	20:00
Air Blank	0.000	20:01
Diagnostics Check	OK	20:01

Data Download Date/Time: 12/18/2017 17:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: SILVERSTEIN,CHRISTINA

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 11/21/2017

Date or Time Changed: No

DRY GAS STANDARD LOT #: 32316080A3

DRY GAS EXPIRY DATE: 01/05/2019

VIOLATION CODE: DUI

ARREST TIME: 01:18

GENDER: Male

OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: DOANE

DATE/TIME OF TEST: 11/26/2017 02:13:30

Diagnostics Check	OK	02:07
Air Blank	0.000	02:07
Control Test	0.080	02:07
Air Blank	0.000	02:08
Subject Sample # 1	0.129	02:09
Breath Volume	3.253	02:09
Air Blank	0.000	02:09
Air Blank	0.000	02:11
Subject Sample # 2	0.131	02:12
Breath Volume	3.121	02:12
Air Blank	0.000	02:12
Control Test	0.080	02:12
Air Blank	0.000	02:13
Diagnostics Check	OK	02:13

Data Download Date/Time: 12/18/2017 17:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: GOSIER,MATTHEW,D

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: GOSIER,MATTHEW

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 11/21/2017

Date or Time Changed: No

DRY GAS STANDARD LOT #: 32316080A3

DRY GAS EXPIRY DATE: 01/05/2019

VIOLATION CODE: DUI

ARREST TIME: 02:04

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: HUSELJIC

DATE/TIME OF TEST: 11/26/2017 02:55:25

Diagnostics Check	OK	02:52	
Air Blank	0.000	02:52	
Control Test	0.080	02:52	
Air Blank	0.000	02:53	
Subject Sample	# 1	02:54	Subject Test Refused
Breath Volume	0.000	02:54	
Air Blank	0.000	02:54	
Control Test	0.080	02:54	
Air Blank	0.000	02:55	
Diagnostics Check	OK	02:55	

Data Download Date/Time: 12/18/2017 17:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: GOSIER, MATTHEW, D

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: MAY, ANDREW

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 11/21/2017

Date or Time Changed: No

DRY GAS STANDARD LOT #: 32316080A3

DRY GAS EXPIRY DATE: 01/05/2019

VIOLATION CODE: DUI

ARREST TIME: 01:43

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:12

SUBJECT'S LAST NAME: JENSEN

DATE/TIME OF TEST: 11/27/2017 02:47:31

Diagnostics Check	OK	02:32	
Air Blank	0.000	02:33	
Control Test	0.079	02:33	
Air Blank	0.000	02:34	
Subject Sample	# 1 0.160	02:37	Volume Not Met
Breath Volume	0.015	02:37	
Air Blank	0.000	02:37	
Air Blank	0.000	02:39	
Subject Sample	# 2 0.159	02:42	
Breath Volume	1.390	02:42	
Air Blank	0.000	02:43	
Air Blank	0.000	02:45	
Subject Sample	# 3 0.158	02:46	
Breath Volume	1.277	02:46	
Air Blank	0.000	02:46	
Control Test	0.079	02:46	
Air Blank	0.000	02:47	
Diagnostics Check	OK	02:47	

Data Download Date/Time: 12/18/2017 17:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: TINDALL,RICHARD,P

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: MAY,ANDREW

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 11/21/2017

Date or Time Changed: No

DRY GAS STANDARD LOT #: 32316080A3

DRY GAS EXPIRY DATE: 01/05/2019

VIOLATION CODE: DUI

ARREST TIME: 23:33

GENDER: Female

OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: RUSSELL

DATE/TIME OF TEST: 11/29/2017 00:18:14

Diagnostics Check	OK	00:11
Air Blank	0.000	00:12
Control Test	0.080	00:12
Air Blank	0.000	00:12
Subject Sample # 1	0.261	00:13
Breath Volume	2.523	00:13
Air Blank	0.000	00:14
Air Blank	0.000	00:16
Subject Sample # 2	0.256	00:16
Breath Volume	2.070	00:16
Air Blank	0.000	00:17
Control Test	0.080	00:17
Air Blank	0.000	00:18
Diagnostics Check	OK	00:18

Data Download Date/Time: 12/18/2017 17:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 11/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:06

SUBJECT'S LAST NAME: SANBORN

DATE/TIME OF TEST: 11/30/2017 04:39:50

Diagnostics Check	OK	04:26	
Air Blank	0.000	04:27	
Control Test	0.081	04:27	
Air Blank	0.000	04:27	
Subject Sample	# 1	04:31	No Sample Provided
Breath Volume	0.000	04:31	
Air Blank	0.000	04:31	
Air Blank	0.000	04:33	
Subject Sample	# 2	0.181	04:34
Breath Volume	1.421	04:34	
Air Blank	0.000	04:35	
Air Blank	0.000	04:36	
Subject Sample	# 3	0.191	04:38
Breath Volume	1.667	04:38	
Air Blank	0.000	04:38	
Control Test	0.080	04:39	
Air Blank	0.000	04:39	
Diagnostics Check	OK	04:39	

Data Download Date/Time: 12/18/2017 17:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HASTINGS,KENNETH,S

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: HASTINGS,KENNETH

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 11/21/2017

Date or Time Changed: No

DRY GAS STANDARD LOT #: 32316080A3

DRY GAS EXPIRY DATE: 01/05/2019

VIOLATION CODE: DUI

ARREST TIME: 04:14

GENDER: Male

OBSERVATION PERIOD BEGAN: 04:25

SUBJECT'S LAST NAME: GOMEZ DEL LEON

DATE/TIME OF TEST: 11/30/2017 04:53:11

Diagnostics Check	OK	04:46
Air Blank	0.000	04:46
Control Test	0.081	04:47
Air Blank	0.000	04:47
Subject Sample # 1	0.188	04:48
Breath Volume	3.062	04:48
Air Blank	0.000	04:49
Air Blank	0.000	04:51
Subject Sample # 2	0.205	04:51
Breath Volume	1.234	04:51
Air Blank	0.000	04:52
Control Test	0.080	04:52
Air Blank	0.000	04:53
Diagnostics Check	OK	04:53

Data Download Date/Time: 12/18/2017 17:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NOTTINGHAM,CHRISTOPHER,R
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: NOTTINGHAM,CHRISTOPHER
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 10/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 19:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:55

SUBJECT'S LAST NAME: PEAT

DATE/TIME OF TEST: 11/01/2017 21:20:18

Diagnostics Check	OK	21:16	
Air Blank	0.000	21:16	
Control Test	0.079	21:17	
Air Blank	0.000	21:17	
Subject Sample	# 1	21:18	Subject Test Refused
Breath Volume	0.000	21:18	
Air Blank	0.000	21:19	
Control Test	0.079	21:19	
Air Blank	0.000	21:20	
Diagnostics Check	OK	21:20	

Data Download Date/Time: 11/29/2017 15:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WINTERS,LANE,A
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: LANE,MARC
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 10/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: CARSWELL

DATE/TIME OF TEST: 11/02/2017 03:42:27

Diagnostics Check	OK	03:35
Air Blank	0.000	03:36
Control Test	0.079	03:36
Air Blank	0.000	03:36
Subject Sample # 1	0.153	03:37
Breath Volume	2.207	03:37
Air Blank	0.000	03:38
Air Blank	0.000	03:40
Subject Sample # 2	0.155	03:40
Breath Volume	1.851	03:40
Air Blank	0.000	03:41
Control Test	0.079	03:41
Air Blank	0.000	03:42
Diagnostics Check	OK	03:42

Data Download Date/Time: 11/29/2017 15:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARQUES,JASON,R
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: POY,ERIC
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 10/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 08:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:50

SUBJECT'S LAST NAME: ONEAL

DATE/TIME OF TEST: 11/04/2017 09:17:15

Diagnostics Check	OK	09:11
Air Blank	0.000	09:11
Control Test	0.079	09:11
Air Blank	0.000	09:12
Subject Sample # 1	0.172	09:12
Breath Volume	2.250	09:12
Air Blank	0.000	09:13
Air Blank	0.000	09:15
Subject Sample # 2	0.161	09:15
Breath Volume	2.109	09:15
Air Blank	0.000	09:16
Control Test	0.079	09:16
Air Blank	0.000	09:17
Diagnostics Check	OK	09:17

Data Download Date/Time: 11/29/2017 15:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CINELLI, ADAM, D
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: CINELLI, ADAM
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 10/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 10:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:08

SUBJECT'S LAST NAME: ELLARS

DATE/TIME OF TEST: 11/04/2017 12:36:53

Diagnostics Check	OK	12:30
Air Blank	0.000	12:31
Control Test	0.080	12:31
Air Blank	0.000	12:32
Subject Sample # 1	0.000	12:32
Breath Volume	1.855	12:32
Air Blank	0.000	12:33
Air Blank	0.000	12:35
Subject Sample # 2	0.000	12:35
Breath Volume	2.074	12:35
Air Blank	0.000	12:35
Control Test	0.080	12:36
Air Blank	0.000	12:36
Diagnostics Check	OK	12:36

Data Download Date/Time: 11/29/2017 15:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CINELLI, ADAM, D
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: CINELLI, ADAM
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 10/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 12:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:43

SUBJECT'S LAST NAME: KELLEHER

DATE/TIME OF TEST: 11/08/2017 14:17:01

Diagnostics Check	OK	14:10
Air Blank	0.000	14:11
Control Test	0.080	14:11
Air Blank	0.000	14:12
Subject Sample # 1	0.160	14:12
Breath Volume	3.328	14:12
Air Blank	0.000	14:13
Air Blank	0.000	14:14
Subject Sample # 2	0.158	14:15
Breath Volume	3.113	14:15
Air Blank	0.000	14:16
Control Test	0.079	14:16
Air Blank	0.000	14:16
Diagnostics Check	OK	14:17

Data Download Date/Time: 11/29/2017 15:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCARTHY,KEITH,
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: SCHELL,MATT
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 10/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: VANDERHAM

DATE/TIME OF TEST: 11/09/2017 01:03:35

Diagnostics Check	OK	00:55
Air Blank	0.000	00:56
Control Test	0.080	00:56
Air Blank	0.000	00:57
Subject Sample # 1	0.076	00:58
Breath Volume	1.800	00:58
Air Blank	0.000	00:59
Air Blank	0.000	01:01
Subject Sample # 2	0.082	01:02
Breath Volume	1.718	01:02
Air Blank	0.000	01:02
Control Test	0.079	01:03
Air Blank	0.000	01:03
Diagnostics Check	OK	01:03

Data Download Date/Time: 11/29/2017 15:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NIXON,BARRY,
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: NIXON,BARRY
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 10/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 19:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:16

SUBJECT'S LAST NAME: CASTRO

DATE/TIME OF TEST: 11/10/2017 20:43:10

Diagnostics Check	OK	20:36
Air Blank	0.000	20:37
Control Test	0.080	20:37
Air Blank	0.000	20:38
Subject Sample # 1	0.000	20:38
Breath Volume	2.816	20:38
Air Blank	0.000	20:39
Air Blank	0.000	20:41
Subject Sample # 2	0.000	20:41
Breath Volume	2.402	20:41
Air Blank	0.000	20:42
Control Test	0.080	20:42
Air Blank	0.000	20:43
Diagnostics Check	OK	20:43

Data Download Date/Time: 11/29/2017 15:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREEN, WILLIAM, E
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: GREEN, WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 10/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: WOERTZ

DATE/TIME OF TEST: 11/10/2017 23:43:45

Diagnostics Check	OK	23:31	
Air Blank	0.000	23:31	
Control Test	0.080	23:31	
Air Blank	0.000	23:32	
Subject Sample	# 1 0.050	23:35	Volume Not Met
Breath Volume	1.007	23:35	
Air Blank	0.000	23:36	
Air Blank	0.000	23:37	
Subject Sample	# 2 0.054	23:39	
Breath Volume	1.445	23:39	
Air Blank	0.000	23:39	
Air Blank	0.000	23:41	
Subject Sample	# 3 0.052	23:42	
Breath Volume	1.445	23:42	
Air Blank	0.000	23:42	
Control Test	0.078	23:43	
Air Blank	0.000	23:43	
Diagnostics Check	OK	23:43	

Data Download Date/Time: 11/29/2017 15:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NOTTINGHAM,CHRISTOPHER,R
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: NOTTINGHAM,CHRISTOPHER
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 10/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: LOGAN

DATE/TIME OF TEST: 11/11/2017 01:22:10

Diagnostics Check	OK	01:15
Air Blank	0.000	01:16
Control Test	0.079	01:16
Air Blank	0.000	01:16
Subject Sample # 1	0.119	01:17
Breath Volume	2.703	01:17
Air Blank	0.000	01:18
Air Blank	0.000	01:19
Subject Sample # 2	0.118	01:20
Breath Volume	2.847	01:20
Air Blank	0.000	01:21
Control Test	0.082	01:21
Air Blank	0.000	01:22
Diagnostics Check	OK	01:22

Data Download Date/Time: 11/29/2017 15:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAWALD,PAUL,M
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: RAWALD,PAUL
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 10/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 19:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:10

SUBJECT'S LAST NAME: DONAHUE

DATE/TIME OF TEST: 11/11/2017 20:33:35

Diagnostics Check	OK	20:30	
Air Blank	0.000	20:31	
Control Test	0.079	20:31	
Air Blank	0.000	20:32	
Subject Sample	# 1	20:32	Subject Test Refused
Breath Volume	0.000	20:32	
Air Blank	0.000	20:32	
Control Test	0.079	20:33	
Air Blank	0.000	20:33	
Diagnostics Check	OK	20:33	

Data Download Date/Time: 11/29/2017 15:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,NELSON,X
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: RODRIGUEZ,NELSON
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 10/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:10

SUBJECT'S LAST NAME: MAS

DATE/TIME OF TEST: 11/15/2017 21:33:32

Diagnostics Check	OK	21:30	
Air Blank	0.000	21:31	
Control Test	0.078	21:31	
Air Blank	0.000	21:31	
Subject Sample	# 1	21:32	Subject Test Refused
Breath Volume	0.000	21:32	
Air Blank	0.000	21:32	
Control Test	0.077	21:32	
Air Blank	0.000	21:33	
Diagnostics Check	OK	21:33	

Data Download Date/Time: 11/29/2017 15:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,JASON,G
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: BROWN,JASON
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 10/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: SANTOS

DATE/TIME OF TEST: 11/15/2017 22:52:41

Diagnostics Check	OK	22:46
Air Blank	0.000	22:47
Control Test	0.081	22:47
Air Blank	0.000	22:47
Subject Sample # 1	0.000	22:48
Breath Volume	2.960	22:48
Air Blank	0.000	22:48
Air Blank	0.000	22:50
Subject Sample # 2	0.000	22:51
Breath Volume	3.472	22:51
Air Blank	0.000	22:51
Control Test	0.080	22:52
Air Blank	0.000	22:52
Diagnostics Check	OK	22:52

Data Download Date/Time: 11/29/2017 15:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,JASON,G
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: BROWN,JASON
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 10/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: GIGANTE

DATE/TIME OF TEST: 11/16/2017 02:34:52

Diagnostics Check	OK	02:28
Air Blank	0.000	02:28
Control Test	0.079	02:29
Air Blank	0.000	02:29
Subject Sample # 1	0.196	02:30
Breath Volume	1.890	02:30
Air Blank	0.000	02:30
Air Blank	0.000	02:32
Subject Sample # 2	0.204	02:33
Breath Volume	2.128	02:33
Air Blank	0.000	02:33
Control Test	0.079	02:34
Air Blank	0.000	02:34
Diagnostics Check	OK	02:34

Data Download Date/Time: 11/29/2017 15:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRIOLAT,CASEY,J
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: BRIOLAT,CASEY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 10/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:12

SUBJECT'S LAST NAME: BERND

DATE/TIME OF TEST: 11/20/2017 03:40:05

Diagnostics Check	OK	03:33
Air Blank	0.000	03:33
Control Test	0.080	03:33
Air Blank	0.000	03:34
Subject Sample # 1	0.147	03:35
Breath Volume	3.503	03:35
Air Blank	0.000	03:35
Air Blank	0.000	03:37
Subject Sample # 2	0.138	03:38
Breath Volume	3.261	03:38
Air Blank	0.000	03:39
Control Test	0.080	03:39
Air Blank	0.000	03:40
Diagnostics Check	OK	03:40

Data Download Date/Time: 11/29/2017 15:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRIOLAT,CASEY,J
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: BRIOLAT,CASEY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 10/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:13

SUBJECT'S LAST NAME: MONSALVE

DATE/TIME OF TEST: 11/23/2017 05:40:57

Diagnostics Check	OK	05:34
Air Blank	0.000	05:34
Control Test	0.080	05:35
Air Blank	0.000	05:35
Subject Sample # 1	0.072	05:36
Breath Volume	2.812	05:36
Air Blank	0.000	05:36
Air Blank	0.000	05:38
Subject Sample # 2	0.071	05:39
Breath Volume	2.617	05:39
Air Blank	0.000	05:40
Control Test	0.079	05:40
Air Blank	0.000	05:40
Diagnostics Check	OK	05:40

Data Download Date/Time: 11/29/2017 15:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,NELSON,X
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: RODRIGUEZ,NELSON
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 10/25/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: KERR

DATE/TIME OF TEST: 11/25/2017 03:04:15

Diagnostics Check	OK	02:57
Air Blank	0.000	02:58
Control Test	0.080	02:58
Air Blank	0.000	02:58
Subject Sample # 1	0.145	02:59
Breath Volume	1.964	02:59
Air Blank	0.000	03:00
Air Blank	0.000	03:02
Subject Sample # 2	0.150	03:02
Breath Volume	2.906	03:02
Air Blank	0.000	03:03
Control Test	0.078	03:03
Air Blank	0.000	03:04
Diagnostics Check	OK	03:04

Data Download Date/Time: 11/29/2017 15:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,NELSON,X
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: RODRIGUEZ,NELSON
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 10/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: KERR

DATE/TIME OF TEST: 11/25/2017 03:16:53

Diagnostics Check	OK	03:10
Air Blank	0.000	03:11
Control Test	0.078	03:11
Air Blank	0.000	03:12
Subject Sample # 1	0.142	03:12
Breath Volume	2.328	03:12
Air Blank	0.000	03:13
Air Blank	0.000	03:14
Subject Sample # 2	0.143	03:15
Breath Volume	1.406	03:15
Air Blank	0.000	03:15
Control Test	0.080	03:16
Air Blank	0.000	03:16
Diagnostics Check	OK	03:16

Data Download Date/Time: 11/29/2017 15:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANE,MARC,T
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: LANE,MARC
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 10/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:32

SUBJECT'S LAST NAME: ROYALL

DATE/TIME OF TEST: 11/25/2017 04:59:32

Diagnostics Check	OK	04:53
Air Blank	0.000	04:53
Control Test	0.080	04:53
Air Blank	0.000	04:54
Subject Sample # 1	0.169	04:55
Breath Volume	1.191	04:55
Air Blank	0.000	04:55
Air Blank	0.000	04:57
Subject Sample # 2	0.177	04:58
Breath Volume	2.550	04:58
Air Blank	0.000	04:58
Control Test	0.079	04:58
Air Blank	0.000	04:59
Diagnostics Check	OK	04:59

Data Download Date/Time: 11/29/2017 15:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLYTH,JACOB,A
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: BLYTH,JACOB
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 10/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: MCELWAIN

DATE/TIME OF TEST: 11/26/2017 01:26:28

Diagnostics Check	OK	01:20
Air Blank	0.000	01:20
Control Test	0.078	01:21
Air Blank	0.000	01:21
Subject Sample # 1	0.181	01:21
Breath Volume	1.867	01:21
Air Blank	0.000	01:22
Air Blank	0.000	01:24
Subject Sample # 2	0.184	01:24
Breath Volume	1.386	01:24
Air Blank	0.000	01:25
Control Test	0.079	01:25
Air Blank	0.000	01:26
Diagnostics Check	OK	01:26

Data Download Date/Time: 11/29/2017 15:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NIXON,BARRY,
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: NIXON,BARRY
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 10/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:17

SUBJECT'S LAST NAME: CHAUNCEY

DATE/TIME OF TEST: 11/26/2017 02:43:17

Diagnostics Check	OK	02:37
Air Blank	0.000	02:37
Control Test	0.080	02:38
Air Blank	0.000	02:38
Subject Sample # 1	0.000	02:39
Breath Volume	1.769	02:39
Air Blank	0.000	02:39
Air Blank	0.000	02:41
Subject Sample # 2	0.000	02:41
Breath Volume	2.031	02:41
Air Blank	0.000	02:42
Control Test	0.080	02:42
Air Blank	0.000	02:43
Diagnostics Check	OK	02:43

Data Download Date/Time: 11/29/2017 15:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NIXON,BARRY,
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: NIXON,BARRY
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 10/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:35

SUBJECT'S LAST NAME: GACOVINO

DATE/TIME OF TEST: 11/26/2017 05:01:58

Diagnostics Check	OK	04:55
Air Blank	0.000	04:55
Control Test	0.080	04:56
Air Blank	0.000	04:56
Subject Sample # 1	0.167	04:57
Breath Volume	2.164	04:57
Air Blank	0.000	04:57
Air Blank	0.000	04:59
Subject Sample # 2	0.156	05:00
Breath Volume	1.335	05:00
Air Blank	0.000	05:01
Control Test	0.079	05:01
Air Blank	0.000	05:01
Diagnostics Check	OK	05:01

Data Download Date/Time: 11/29/2017 15:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANE,MARC,T
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: JOHANSEN,LLOYD
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 10/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:52

SUBJECT'S LAST NAME: CONKLIN

DATE/TIME OF TEST: 11/27/2017 03:27:00

Diagnostics Check	OK	03:12	
Air Blank	0.000	03:12	
Control Test	0.080	03:13	
Air Blank	0.000	03:13	
Subject Sample	# 1 0.053	03:16	
Breath Volume	1.703	03:16	
Air Blank	0.000	03:17	
Air Blank	0.000	03:19	
Subject Sample	# 2 0.059	03:22	Volume Not Met
Breath Volume	0.320	03:22	
Air Blank	0.000	03:22	
Air Blank	0.000	03:24	
Subject Sample	# 3 0.060	03:25	
Breath Volume	2.636	03:25	
Air Blank	0.000	03:26	
Control Test	0.078	03:26	
Air Blank	0.000	03:26	
Diagnostics Check	OK	03:27	

Data Download Date/Time: 11/29/2017 15:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NIXON,BARRY,
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: NIXON,BARRY
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 11/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 19:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:17

SUBJECT'S LAST NAME: VOSBURGH

DATE/TIME OF TEST: 11/30/2017 20:44:18

Diagnostics Check	OK	20:37
Air Blank	0.000	20:38
Control Test	0.081	20:38
Air Blank	0.000	20:39
Subject Sample # 1	0.000	20:39
Breath Volume	3.742	20:39
Air Blank	0.000	20:40
Air Blank	0.000	20:42
Subject Sample # 2	0.000	20:42
Breath Volume	2.187	20:42
Air Blank	0.000	20:43
Control Test	0.080	20:43
Air Blank	0.000	20:44
Diagnostics Check	OK	20:44

Data Download Date/Time: 12/27/2017 16:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001082

INSTRUMENT REGISTERED TO: PLANTATION PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROSKIE,OFC. J.,
BREATH TEST OPERATOR AGENCY: Plantation PD
ARRESTING OFFICER: FARQUHARSON,OFC. K.
ARRESTING OFFICER AGENCY: Plantation PD
LAST AGENCY INSPECTION: 10/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A2
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:11

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 11/19/2017 04:32:42

Diagnostics Check	OK	04:31	
Air Blank	0.000	04:31	
Control Test	0.000	04:32	Control Outside Tolerance
Air Blank	0.000	04:32	

Data Download Date/Time: 11/23/2017 14:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001082

INSTRUMENT REGISTERED TO: PLANTATION PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BROSKIE,OFC. J.,

BREATH TEST OPERATOR AGENCY: Plantation PD

ARRESTING OFFICER: FARQUHARSON,OFC. K.

ARRESTING OFFICER AGENCY: Plantation PD

LAST AGENCY INSPECTION: 10/19/2017

Date or Time Changed: No

DRY GAS STANDARD LOT #: 14816080A2

DRY GAS EXPIRY DATE: 07/05/2018

VIOLATION CODE: DUI

ARREST TIME: 04:00

GENDER: Female

OBSERVATION PERIOD BEGAN: 04:11

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 11/19/2017 04:36:24

Diagnostics Check	OK	04:35	
Air Blank	0.000	04:35	
Control Test	0.000	04:35	Control Outside Tolerance
Air Blank	0.000	04:36	

Data Download Date/Time: 11/23/2017 14:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001082

INSTRUMENT REGISTERED TO: PLANTATION PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROSKIE,OFC. J.,
BREATH TEST OPERATOR AGENCY: Plantation PD
ARRESTING OFFICER: FARQUHARSON,OFC. K.
ARRESTING OFFICER AGENCY: Plantation PD
LAST AGENCY INSPECTION: 10/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A2
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:11

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 11/19/2017 04:39:57

Diagnostics Check	OK	04:38	
Air Blank	0.000	04:39	
Control Test	0.000	04:39	Control Outside Tolerance
Air Blank	0.000	04:39	

Data Download Date/Time: 11/23/2017 14:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001082

INSTRUMENT REGISTERED TO: PLANTATION PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BROSKIE,OFC. J.,

BREATH TEST OPERATOR AGENCY: Plantation PD

ARRESTING OFFICER: FARQUHARSON,OFC. K.

ARRESTING OFFICER AGENCY: Plantation PD

LAST AGENCY INSPECTION: 10/19/2017

Date or Time Changed: No

DRY GAS STANDARD LOT #: 14816080A2

DRY GAS EXPIRY DATE: 07/05/2018

VIOLATION CODE: DUI

ARREST TIME: 04:00

GENDER: Female

OBSERVATION PERIOD BEGAN: 04:11

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 11/19/2017 04:43:27

Diagnostics Check	OK	04:42	
Air Blank	0.000	04:42	
Control Test	0.000	04:42	Control Outside Tolerance
Air Blank	0.000	04:43	

Data Download Date/Time: 11/23/2017 14:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001082

INSTRUMENT REGISTERED TO: PLANTATION PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROSKIE,OFC. J.,
BREATH TEST OPERATOR AGENCY: Plantation PD
ARRESTING OFFICER: FARQUHARSON,OFC. K.
ARRESTING OFFICER AGENCY: Plantation PD
LAST AGENCY INSPECTION: 10/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A2
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:11

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 11/19/2017 04:53:20

Diagnostics Check	OK	04:52	
Air Blank	0.000	04:52	
Control Test	0.000	04:52	Control Outside Tolerance
Air Blank	0.000	04:53	

Data Download Date/Time: 11/23/2017 14:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001082

INSTRUMENT REGISTERED TO: PLANTATION PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BEAUREGARD,JEFF,
BREATH TEST OPERATOR AGENCY: Plantation PD
ARRESTING OFFICER: BEAUREGARD,JEFF
ARRESTING OFFICER AGENCY: Plantation PD
LAST AGENCY INSPECTION: 11/23/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A2
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:17

SUBJECT'S LAST NAME: AVALOS

DATE/TIME OF TEST: 11/25/2017 02:53:16

Diagnostics Check	OK	02:39	
Air Blank	0.000	02:39	
Control Test	0.077	02:40	
Air Blank	0.000	02:40	
Subject Sample	# 1 0.081	02:43	Volume Not Met
Breath Volume	0.207	02:43	
Air Blank	0.000	02:44	
Air Blank	0.000	02:46	
Subject Sample	# 2 0.090	02:47	
Breath Volume	1.804	02:47	
Air Blank	0.000	02:48	
Air Blank	0.000	02:50	
Subject Sample	# 3 0.094	02:51	
Breath Volume	1.917	02:51	
Air Blank	0.000	02:52	
Control Test	0.076	02:52	
Air Blank	0.000	02:53	
Diagnostics Check	OK	02:53	

Data Download Date/Time: 12/28/2017 19:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBER,JOSEPH,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: THOMPSON,M.E.
ARRESTING OFFICER AGENCY: Atlantic Beach PD
LAST AGENCY INSPECTION: 10/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:11

SUBJECT'S LAST NAME: WATTS

DATE/TIME OF TEST: 11/01/2017 01:34:49

Diagnostics Check	OK	01:31	
Air Blank	0.000	01:32	
Control Test	0.080	01:32	
Air Blank	0.000	01:33	
Subject Sample	# 1	01:33	Subject Test Refused
Breath Volume	0.000	01:33	
Air Blank	0.000	01:33	
Control Test	0.080	01:34	
Air Blank	0.000	01:34	
Diagnostics Check	OK	01:34	

Data Download Date/Time: 11/01/2017 12:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WBER,JOSEPH,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FILBERT,E.D.
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 10/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:43

SUBJECT'S LAST NAME: SHERIFFS

DATE/TIME OF TEST: 11/01/2017 03:15:54

Diagnostics Check	OK	03:09	
Air Blank	0.000	03:09	
Control Test	0.081	03:09	
Air Blank	0.000	03:10	
Subject Sample # 1	0.083	03:11	
Breath Volume	1.773	03:11	
Air Blank	0.000	03:12	
Air Blank	0.000	03:14	
		03:14	RFI Detect
Air Blank	0.000	03:15	
Control Test	0.080	03:15	
Air Blank	0.000	03:15	
Diagnostics Check	OK	03:15	

Data Download Date/Time: 11/01/2017 12:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBER,JOSEPH,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FILBERT,E.D.
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 10/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:43

SUBJECT'S LAST NAME: SHERIFFS

DATE/TIME OF TEST: 11/01/2017 03:25:23

Diagnostics Check	OK	03:16
Air Blank	0.000	03:17
Control Test	0.080	03:17
Air Blank	0.000	03:18
Subject Sample # 1	0.077	03:20
Breath Volume	1.558	03:20
Air Blank	0.000	03:20
Air Blank	0.000	03:22
Subject Sample # 2	0.078	03:23
Breath Volume	2.289	03:23
Air Blank	0.000	03:24
Control Test	0.079	03:24
Air Blank	0.000	03:25
Diagnostics Check	OK	03:25

Data Download Date/Time: 11/01/2017 12:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBER,JOSEPH,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: KEEN,C.E.
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 10/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:52

SUBJECT'S LAST NAME: CADENA

DATE/TIME OF TEST: 11/01/2017 04:16:41

Diagnostics Check	OK	04:13	
Air Blank	0.000	04:13	
Control Test	0.080	04:13	
Air Blank	0.000	04:14	
Subject Sample	# 1	04:15	Subject Test Refused
Breath Volume	0.000	04:15	
Air Blank	0.000	04:15	
Control Test	0.080	04:16	
Air Blank	0.000	04:16	
Diagnostics Check	OK	04:16	

Data Download Date/Time: 11/01/2017 12:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENDRIX,D,K
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MANNING,N
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:36

SUBJECT'S LAST NAME: TALLMAN

DATE/TIME OF TEST: 11/08/2017 07:03:16

Diagnostics Check	OK	06:57
Air Blank	0.000	06:57
Control Test	0.080	06:57
Air Blank	0.000	06:58
Subject Sample # 1	0.145	06:58
Breath Volume	1.304	06:58
Air Blank	0.000	06:59
Air Blank	0.000	07:01
Subject Sample # 2	0.149	07:01
Breath Volume	2.910	07:01
Air Blank	0.000	07:02
Control Test	0.079	07:02
Air Blank	0.000	07:03
Diagnostics Check	OK	07:03

Data Download Date/Time: 11/17/2017 13:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CARLSON,P
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 11/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: MATHEWS

DATE/TIME OF TEST: 11/09/2017 03:31:58

Diagnostics Check	OK	03:25
Air Blank	0.000	03:26
Control Test	0.079	03:26
Air Blank	0.000	03:26
Subject Sample # 1	0.080	03:27
Breath Volume	3.480	03:27
Air Blank	0.000	03:28
Air Blank	0.000	03:30
Subject Sample # 2	0.082	03:30
Breath Volume	3.980	03:30
Air Blank	0.000	03:31
Control Test	0.078	03:31
Air Blank	0.000	03:31
Diagnostics Check	OK	03:31

Data Download Date/Time: 11/17/2017 13:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENDRIX,D,K
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CAMPBELL,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 11:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:30

SUBJECT'S LAST NAME: LAUGHREN

DATE/TIME OF TEST: 11/09/2017 13:58:10

Diagnostics Check	OK	13:51
Air Blank	0.000	13:52
Control Test	0.079	13:52
Air Blank	0.000	13:52
Subject Sample # 1	0.000	13:53
Breath Volume	1.175	13:53
Air Blank	0.000	13:54
Air Blank	0.000	13:56
Subject Sample # 2	0.000	13:56
Breath Volume	2.167	13:56
Air Blank	0.000	13:57
Control Test	0.079	13:57
Air Blank	0.000	13:58
Diagnostics Check	OK	13:58

Data Download Date/Time: 11/17/2017 13:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FARLEY,J
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 11/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:38

SUBJECT'S LAST NAME: HORTON

DATE/TIME OF TEST: 11/10/2017 00:04:55

Diagnostics Check	OK	23:58
Air Blank	0.000	23:59
Control Test	0.080	23:59
Air Blank	0.000	23:59
Subject Sample # 1	0.217	00:00
Breath Volume	2.003	00:00
Air Blank	0.000	00:00
Air Blank	0.000	00:02
Subject Sample # 2	0.213	00:03
Breath Volume	2.320	00:03
Air Blank	0.000	00:04
Control Test	0.080	00:04
Air Blank	0.000	00:04
Diagnostics Check	OK	00:04

Data Download Date/Time: 11/17/2017 13:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CARLSON,P-
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 11/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:26

SUBJECT'S LAST NAME: ROBINSON

DATE/TIME OF TEST: 11/10/2017 01:50:00

Diagnostics Check	OK	01:46	
Air Blank	0.000	01:47	
Control Test	0.079	01:47	
Air Blank	0.000	01:48	
Subject Sample	# 1	01:48	Subject Test Refused
Breath Volume	0.000	01:48	
Air Blank	0.000	01:49	
Control Test	0.080	01:49	
Air Blank	0.000	01:49	
Diagnostics Check	OK	01:50	

Data Download Date/Time: 11/17/2017 13:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FARLEY,J
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 11/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:46

SUBJECT'S LAST NAME: WADE

DATE/TIME OF TEST: 11/11/2017 00:13:48

Diagnostics Check	OK	00:07
Air Blank	0.000	00:07
Control Test	0.081	00:07
Air Blank	0.000	00:08
Subject Sample # 1	0.064	00:09
Breath Volume	2.296	00:09
Air Blank	0.000	00:10
Air Blank	0.000	00:11
Subject Sample # 2	0.060	00:12
Breath Volume	2.066	00:12
Air Blank	0.000	00:12
Control Test	0.079	00:13
Air Blank	0.000	00:13
Diagnostics Check	OK	00:13

Data Download Date/Time: 11/17/2017 13:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CARLSON,O
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 11/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: COKER

DATE/TIME OF TEST: 11/11/2017 03:53:48

Diagnostics Check	OK	03:50	
Air Blank	0.000	03:51	
Control Test	0.080	03:51	
Air Blank	0.000	03:52	
Subject Sample	# 1	03:52	Subject Test Refused
Breath Volume	0.000	03:52	
Air Blank	0.000	03:52	
Control Test	0.080	03:53	
Air Blank	0.000	03:53	
Diagnostics Check	OK	03:53	

Data Download Date/Time: 11/17/2017 13:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: THOMAS,J
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 11/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:25

SUBJECT'S LAST NAME: PHILLIPS

DATE/TIME OF TEST: 11/11/2017 04:53:41

Diagnostics Check	OK	04:46
Air Blank	0.000	04:46
Control Test	0.080	04:46
Air Blank	0.000	04:47
Subject Sample # 1	0.074	04:49
Breath Volume	2.195	04:49
Air Blank	0.000	04:49
Air Blank	0.000	04:51
Subject Sample # 2	0.071	04:52
Breath Volume	1.621	04:52
Air Blank	0.000	04:52
Control Test	0.079	04:53
Air Blank	0.000	04:53
Diagnostics Check	OK	04:53

Data Download Date/Time: 11/17/2017 13:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: KEEN,C
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 11/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 18:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: ROZESTRATEN

DATE/TIME OF TEST: 11/11/2017 22:28:23

Diagnostics Check	OK	22:25	
Air Blank	0.000	22:26	
Control Test	0.081	22:26	
Air Blank	0.000	22:26	
Subject Sample	# 1	22:27	Subject Test Refused
Breath Volume	0.000	22:27	
Air Blank	0.000	22:27	
Control Test	0.080	22:27	
Air Blank	0.000	22:28	
Diagnostics Check	OK	22:28	

Data Download Date/Time: 11/17/2017 13:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MOELLER,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: CARTER

DATE/TIME OF TEST: 11/12/2017 00:11:14

Diagnostics Check	OK	00:05
Air Blank	0.000	00:05
Control Test	0.080	00:06
Air Blank	0.000	00:06
Subject Sample # 1	0.082	00:06
Breath Volume	3.683	00:06
Air Blank	0.000	00:07
Air Blank	0.000	00:09
Subject Sample # 2	0.086	00:09
Breath Volume	3.980	00:09
Air Blank	0.000	00:10
Control Test	0.080	00:10
Air Blank	0.000	00:11
Diagnostics Check	OK	00:11

Data Download Date/Time: 11/17/2017 13:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FILBERT,E
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 11/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:12

SUBJECT'S LAST NAME: WOOTEN

DATE/TIME OF TEST: 11/12/2017 02:38:59

Diagnostics Check	OK	02:32
Air Blank	0.000	02:33
Control Test	0.080	02:33
Air Blank	0.000	02:33
Subject Sample # 1	0.000	02:34
Breath Volume	2.097	02:34
Air Blank	0.000	02:35
Air Blank	0.000	02:37
Subject Sample # 2	0.000	02:37
Breath Volume	2.410	02:37
Air Blank	0.000	02:38
Control Test	0.080	02:38
Air Blank	0.000	02:38
Diagnostics Check	OK	02:38

Data Download Date/Time: 11/17/2017 13:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FISETTE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:57

SUBJECT'S LAST NAME: BANKS

DATE/TIME OF TEST: 11/12/2017 03:24:14

Diagnostics Check	OK	03:17
Air Blank	0.000	03:18
Control Test	0.081	03:18
Air Blank	0.000	03:18
Subject Sample # 1	0.252	03:19
Breath Volume	1.621	03:19
Air Blank	0.000	03:19
Air Blank	0.000	03:21
Subject Sample # 2	0.241	03:22
Breath Volume	1.496	03:22
Air Blank	0.000	03:23
Control Test	0.080	03:23
Air Blank	0.000	03:24
Diagnostics Check	OK	03:24

Data Download Date/Time: 11/17/2017 13:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHITE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:54

SUBJECT'S LAST NAME: SIFUNTES

DATE/TIME OF TEST: 11/12/2017 04:20:40

Diagnostics Check	OK	04:14
Air Blank	0.000	04:14
Control Test	0.080	04:15
Air Blank	0.000	04:15
Subject Sample # 1	0.085	04:16
Breath Volume	2.722	04:16
Air Blank	0.000	04:16
Air Blank	0.000	04:18
Subject Sample # 2	0.082	04:19
Breath Volume	3.328	04:19
Air Blank	0.000	04:19
Control Test	0.080	04:20
Air Blank	0.000	04:20
Diagnostics Check	OK	04:20

Data Download Date/Time: 11/17/2017 13:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS, CHRISTOPHER, D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MOELLER, R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 17:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:48

SUBJECT'S LAST NAME: SAVAGE

DATE/TIME OF TEST: 11/12/2017 19:14:58

Diagnostics Check	OK	19:08
Air Blank	0.000	19:09
Control Test	0.079	19:09
Air Blank	0.000	19:10
Subject Sample # 1	0.072	19:10
Breath Volume	2.562	19:10
Air Blank	0.000	19:11
Air Blank	0.000	19:13
Subject Sample # 2	0.068	19:13
Breath Volume	1.449	19:13
Air Blank	0.000	19:14
Control Test	0.079	19:14
Air Blank	0.000	19:14
Diagnostics Check	OK	19:14

Data Download Date/Time: 11/17/2017 13:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: STANLEY,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:53

SUBJECT'S LAST NAME: SORTO MEJIA

DATE/TIME OF TEST: 11/12/2017 22:16:45

Diagnostics Check	OK	22:13	
Air Blank	0.000	22:14	
Control Test	0.079	22:14	
Air Blank	0.000	22:15	
Subject Sample	# 1	22:15	Subject Test Refused
Breath Volume	0.000	22:15	
Air Blank	0.000	22:15	
Control Test	0.080	22:16	
Air Blank	0.000	22:16	
Diagnostics Check	OK	22:16	

Data Download Date/Time: 11/17/2017 13:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HAYES,D
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 11/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:07

SUBJECT'S LAST NAME: HALILOVIC

DATE/TIME OF TEST: 11/12/2017 23:33:41

Diagnostics Check	OK	23:27
Air Blank	0.000	23:28
Control Test	0.079	23:28
Air Blank	0.000	23:28
Subject Sample # 1	0.090	23:29
Breath Volume	1.429	23:29
Air Blank	0.000	23:29
Air Blank	0.000	23:31
Subject Sample # 2	0.089	23:32
Breath Volume	1.531	23:32
Air Blank	0.000	23:32
Control Test	0.080	23:33
Air Blank	0.000	23:33
Diagnostics Check	OK	23:33

Data Download Date/Time: 11/17/2017 13:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ROSSIGNOL,H
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 11/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:38

SUBJECT'S LAST NAME: PALENZUELA

DATE/TIME OF TEST: 11/13/2017 02:05:02

Diagnostics Check	OK	01:58
Air Blank	0.000	01:58
Control Test	0.080	01:59
Air Blank	0.000	01:59
Subject Sample # 1	0.088	02:00
Breath Volume	2.597	02:00
Air Blank	0.000	02:01
Air Blank	0.000	02:03
Subject Sample # 2	0.089	02:03
Breath Volume	1.652	02:03
Air Blank	0.000	02:04
Control Test	0.080	02:04
Air Blank	0.000	02:04
Diagnostics Check	OK	02:05

Data Download Date/Time: 11/17/2017 13:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CAMACHO,K
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 11/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:23

SUBJECT'S LAST NAME: HOFFMAN

DATE/TIME OF TEST: 11/13/2017 02:46:55

Diagnostics Check	OK	02:44	
Air Blank	0.000	02:44	
Control Test	0.080	02:44	
Air Blank	0.000	02:45	
Subject Sample	# 1	02:45	Subject Test Refused
Breath Volume	0.000	02:45	
Air Blank	0.000	02:46	
Control Test	0.080	02:46	
Air Blank	0.000	02:46	
Diagnostics Check	OK	02:46	

Data Download Date/Time: 11/17/2017 13:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: GIBSON,D
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 11/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:48

SUBJECT'S LAST NAME: FERRELL

DATE/TIME OF TEST: 11/13/2017 03:11:31

Diagnostics Check	OK	03:08	
Air Blank	0.000	03:09	
Control Test	0.080	03:09	
Air Blank	0.000	03:09	
Subject Sample	# 1	03:10	Subject Test Refused
Breath Volume	0.000	03:10	
Air Blank	0.000	03:10	
Control Test	0.080	03:10	
Air Blank	0.000	03:11	
Diagnostics Check	OK	03:11	

Data Download Date/Time: 11/17/2017 13:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: STANLEY,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:34

SUBJECT'S LAST NAME: GARMON

DATE/TIME OF TEST: 11/13/2017 05:01:21

Diagnostics Check	OK	04:55
Air Blank	0.000	04:55
Control Test	0.080	04:55
Air Blank	0.000	04:56
Subject Sample # 1	0.156	04:57
Breath Volume	1.941	04:57
Air Blank	0.000	04:57
Air Blank	0.000	04:59
Subject Sample # 2	0.151	04:59
Breath Volume	1.363	04:59
Air Blank	0.000	05:00
Control Test	0.080	05:00
Air Blank	0.000	05:01
Diagnostics Check	OK	05:01

Data Download Date/Time: 11/17/2017 13:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: STANLEY,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:27

SUBJECT'S LAST NAME: LOPEZ HERNANDEZ

DATE/TIME OF TEST: 11/15/2017 03:50:17

Diagnostics Check	OK	03:47	
Air Blank	0.000	03:47	
Control Test	0.081	03:48	
Air Blank	0.000	03:48	
Subject Sample	# 1	03:48	Subject Test Refused
Breath Volume	0.000	03:48	
Air Blank	0.000	03:49	
Control Test	0.081	03:49	
Air Blank	0.000	03:50	
Diagnostics Check	OK	03:50	

Data Download Date/Time: 11/17/2017 13:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MOELLER,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:08

SUBJECT'S LAST NAME: FAIRCLOTH

DATE/TIME OF TEST: 11/15/2017 23:34:43

Diagnostics Check	OK	23:28
Air Blank	0.000	23:29
Control Test	0.081	23:29
Air Blank	0.000	23:29
Subject Sample # 1	0.131	23:30
Breath Volume	2.515	23:30
Air Blank	0.000	23:30
Air Blank	0.000	23:32
Subject Sample # 2	0.132	23:33
Breath Volume	2.589	23:33
Air Blank	0.000	23:33
Control Test	0.080	23:34
Air Blank	0.000	23:34
Diagnostics Check	OK	23:34

Data Download Date/Time: 11/17/2017 13:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: STANLEY,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:12

SUBJECT'S LAST NAME: WASHINGTON

DATE/TIME OF TEST: 11/16/2017 00:38:37

Diagnostics Check	OK	00:32
Air Blank	0.000	00:32
Control Test	0.080	00:33
Air Blank	0.000	00:33
Subject Sample # 1	0.160	00:34
Breath Volume	1.792	00:34
Air Blank	0.000	00:34
Air Blank	0.000	00:36
Subject Sample # 2	0.162	00:37
Breath Volume	1.730	00:37
Air Blank	0.000	00:37
Control Test	0.080	00:38
Air Blank	0.000	00:38
Diagnostics Check	OK	00:38

Data Download Date/Time: 11/17/2017 13:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HOOVER,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:53

SUBJECT'S LAST NAME: PEARSON

DATE/TIME OF TEST: 11/16/2017 02:16:33

Diagnostics Check	OK	02:13	
Air Blank	0.000	02:13	
Control Test	0.080	02:14	
Air Blank	0.000	02:14	
Subject Sample	# 1	02:15	Subject Test Refused
Breath Volume	0.000	02:15	
Air Blank	0.000	02:15	
Control Test	0.080	02:15	
Air Blank	0.000	02:16	
Diagnostics Check	OK	02:16	

Data Download Date/Time: 11/17/2017 13:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CAMACHO,K
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 11/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:07

SUBJECT'S LAST NAME: TODD

DATE/TIME OF TEST: 11/16/2017 03:36:53

Diagnostics Check	OK	03:27
Air Blank	0.000	03:27
Control Test	0.080	03:27
Air Blank	0.000	03:28
Subject Sample # 1	0.243	03:28
Breath Volume	2.050	03:28
Air Blank	0.000	03:29
Air Blank	0.000	03:31
Subject Sample # 2	0.210	03:32
Breath Volume	1.476	03:32
Air Blank	0.000	03:32
Air Blank	0.000	03:34
Subject Sample # 3	0.227	03:35
Breath Volume	2.691	03:35
Air Blank	0.000	03:35
Control Test	0.080	03:36
Air Blank	0.000	03:36
Diagnostics Check	OK	03:36

Data Download Date/Time: 11/17/2017 13:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURNS,D,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CARLSON,P
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 11/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:28

SUBJECT'S LAST NAME: NEVERSON

DATE/TIME OF TEST: 11/17/2017 03:58:03

Diagnostics Check	OK	03:49
Air Blank	0.000	03:49
Control Test	0.080	03:49
Air Blank	0.000	03:50
Subject Sample # 1	0.143	03:53
Breath Volume	1.550	03:53
Air Blank	0.000	03:53
Air Blank	0.000	03:55
Subject Sample # 2	0.148	03:56
Breath Volume	2.085	03:56
Air Blank	0.000	03:57
Control Test	0.079	03:57
Air Blank	0.000	03:57
Diagnostics Check	OK	03:58

Data Download Date/Time: 11/17/2017 13:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BRYANT,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:06

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 11/18/2017 01:32:30

Diagnostics Check	OK	01:26
Air Blank	0.000	01:27
Control Test	0.080	01:27
Air Blank	0.000	01:27
Subject Sample # 1	0.106	01:28
Breath Volume	2.152	01:28
Air Blank	0.000	01:28
Air Blank	0.000	01:30
Subject Sample # 2	0.103	01:30
Breath Volume	2.542	01:30
Air Blank	0.000	01:31
Control Test	0.079	01:31
Air Blank	0.000	01:32
Diagnostics Check	OK	01:32

Data Download Date/Time: 12/11/2017 15:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SIMMONS,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:22

SUBJECT'S LAST NAME: FOSKEY

DATE/TIME OF TEST: 11/18/2017 03:55:32

Diagnostics Check	OK	03:43	
Air Blank	0.000	03:43	
Control Test	0.079	03:43	
Air Blank	0.000	03:44	
Subject Sample	# 1 0.206	03:47	Volume Not Met
Breath Volume	0.937	03:47	
Air Blank	0.000	03:48	
Air Blank	0.000	03:50	
Subject Sample	# 2 0.177	03:50	
Breath Volume	1.847	03:50	
Air Blank	0.000	03:51	
Air Blank	0.000	03:53	
Subject Sample	# 3 0.197	03:54	
Breath Volume	1.253	03:54	
Air Blank	0.000	03:54	
Control Test	0.078	03:54	
Air Blank	0.000	03:55	
Diagnostics Check	OK	03:55	

Data Download Date/Time: 12/11/2017 15:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CARLSON,P
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:12

SUBJECT'S LAST NAME: COKER

DATE/TIME OF TEST: 11/18/2017 23:39:01

Diagnostics Check	OK	23:32
Air Blank	0.000	23:33
Control Test	0.080	23:33
Air Blank	0.000	23:34
Subject Sample # 1	0.104	23:34
Breath Volume	2.171	23:34
Air Blank	0.000	23:35
Air Blank	0.000	23:37
Subject Sample # 2	0.101	23:37
Breath Volume	2.207	23:37
Air Blank	0.000	23:38
Control Test	0.079	23:38
Air Blank	0.000	23:38
Diagnostics Check	OK	23:39

Data Download Date/Time: 12/11/2017 15:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BROWN,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:38

SUBJECT'S LAST NAME: FARMER JR

DATE/TIME OF TEST: 11/19/2017 03:02:42

Diagnostics Check	OK	02:59	
Air Blank	0.000	03:00	
Control Test	0.080	03:00	
Air Blank	0.000	03:01	
Subject Sample	# 1	03:01	Subject Test Refused
Breath Volume	0.000	03:01	
Air Blank	0.000	03:01	
Control Test	0.081	03:02	
Air Blank	0.000	03:02	
Diagnostics Check	OK	03:02	

Data Download Date/Time: 12/11/2017 15:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CAMACHO,K
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:18

SUBJECT'S LAST NAME: MALEVAN

DATE/TIME OF TEST: 11/22/2017 01:45:33

Diagnostics Check	OK	01:38
Air Blank	0.000	01:39
Control Test	0.081	01:39
Air Blank	0.000	01:40
Subject Sample # 1	0.074	01:40
Breath Volume	1.839	01:40
Air Blank	0.000	01:41
Air Blank	0.000	01:43
Subject Sample # 2	0.074	01:44
Breath Volume	2.199	01:44
Air Blank	0.000	01:44
Control Test	0.080	01:44
Air Blank	0.000	01:45
Diagnostics Check	OK	01:45

Data Download Date/Time: 12/11/2017 15:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DOUGHFMAN,D
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:45

SUBJECT'S LAST NAME: ALIJA

DATE/TIME OF TEST: 11/22/2017 05:12:26

Diagnostics Check	OK	05:05
Air Blank	0.000	05:06
Control Test	0.080	05:06
Air Blank	0.000	05:07
Subject Sample # 1	0.000	05:08
Breath Volume	2.476	05:08
Air Blank	0.000	05:08
Air Blank	0.000	05:10
Subject Sample # 2	0.000	05:11
Breath Volume	3.101	05:11
Air Blank	0.000	05:11
Control Test	0.080	05:11
Air Blank	0.000	05:12
Diagnostics Check	OK	05:12

Data Download Date/Time: 12/11/2017 15:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FISETTE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:06

SUBJECT'S LAST NAME: ROSENLUND

DATE/TIME OF TEST: 11/23/2017 00:30:16

Diagnostics Check	OK	00:26	
Air Blank	0.000	00:27	
Control Test	0.080	00:27	
Air Blank	0.000	00:28	
Subject Sample # 1	0.102	00:28	
Breath Volume	4.546	00:28	
Air Blank		00:28	RFI Detect
Air Blank	0.000	00:29	
Control Test	0.080	00:29	
Air Blank	0.000	00:30	
Diagnostics Check	OK	00:30	

Data Download Date/Time: 12/11/2017 15:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FISETTE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:06

SUBJECT'S LAST NAME: ROSENLUND

DATE/TIME OF TEST: 11/23/2017 00:38:44

Diagnostics Check	OK	00:32
Air Blank	0.000	00:32
Control Test	0.079	00:33
Air Blank	0.000	00:33
Subject Sample # 1	0.100	00:34
Breath Volume	3.691	00:34
Air Blank	0.000	00:34
Air Blank	0.000	00:36
Subject Sample # 2	0.096	00:37
Breath Volume	2.906	00:37
Air Blank	0.000	00:37
Control Test	0.079	00:38
Air Blank	0.000	00:38
Diagnostics Check	OK	00:38

Data Download Date/Time: 12/11/2017 15:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MOELLER,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:38

SUBJECT'S LAST NAME: MCKENZIE

DATE/TIME OF TEST: 11/23/2017 02:07:51

Diagnostics Check	OK	02:05	
Air Blank	0.000	02:05	
Control Test	0.080	02:05	
Air Blank	0.000	02:06	
Subject Sample	# 1	02:06	Subject Test Refused
Breath Volume	0.000	02:06	
Air Blank	0.000	02:06	
Control Test	0.079	02:07	
Air Blank	0.000	02:07	
Diagnostics Check	OK	02:07	

Data Download Date/Time: 12/11/2017 15:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FILBERT,E
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:44

SUBJECT'S LAST NAME: EVANS

DATE/TIME OF TEST: 11/23/2017 04:08:14

Diagnostics Check	OK	04:05	
Air Blank	0.000	04:05	
Control Test	0.080	04:06	
Air Blank	0.000	04:06	
Subject Sample	# 1	04:06	Subject Test Refused
Breath Volume	0.000	04:06	
Air Blank	0.000	04:07	
Control Test	0.081	04:07	
Air Blank	0.000	04:08	
Diagnostics Check	OK	04:08	

Data Download Date/Time: 12/11/2017 15:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MOELLER,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: NAPIER

DATE/TIME OF TEST: 11/24/2017 01:51:51

Diagnostics Check	OK	01:45
Air Blank	0.000	01:46
Control Test	0.080	01:46
Air Blank	0.000	01:47
Subject Sample # 1	0.090	01:47
Breath Volume	2.210	01:47
Air Blank	0.000	01:48
Air Blank	0.000	01:49
Subject Sample # 2	0.090	01:50
Breath Volume	3.000	01:50
Air Blank	0.000	01:50
Control Test	0.080	01:51
Air Blank	0.000	01:51
Diagnostics Check	OK	01:51

Data Download Date/Time: 12/11/2017 15:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: GARCIA,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:06

SUBJECT'S LAST NAME: GIBSON

DATE/TIME OF TEST: 11/24/2017 03:28:39

Diagnostics Check	OK	03:26	
Air Blank	0.000	03:26	
Control Test	0.081	03:27	
Air Blank	0.000	03:27	
Subject Sample	# 1	03:27	Subject Test Refused
Breath Volume	0.000	03:27	
Air Blank		03:28	RFI Detect
Air Blank	0.000	03:28	

Data Download Date/Time: 12/11/2017 15:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: GARCIA,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:06

SUBJECT'S LAST NAME: GIBSON

DATE/TIME OF TEST: 11/24/2017 03:32:54

Diagnostics Check	OK	03:29	
Air Blank	0.000	03:30	
Control Test	0.080	03:30	
Air Blank	0.000	03:31	
Subject Sample	# 1	03:31	Subject Test Refused
Breath Volume	0.000	03:31	
Air Blank	0.000	03:31	
Control Test	0.080	03:32	
Air Blank	0.000	03:32	
Diagnostics Check	OK	03:32	

Data Download Date/Time: 12/11/2017 15:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BYRNE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: JAZEXHI

DATE/TIME OF TEST: 11/27/2017 01:45:01

Diagnostics Check	OK	01:42	
Air Blank	0.000	01:42	
Control Test	0.080	01:42	
Air Blank	0.000	01:43	
Subject Sample	# 1	01:43	Subject Test Refused
Breath Volume	0.000	01:43	
Air Blank	0.000	01:44	
Control Test	0.080	01:44	
Air Blank	0.000	01:44	
Diagnostics Check	OK	01:45	

Data Download Date/Time: 12/11/2017 15:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SORRELLS,M
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:57

SUBJECT'S LAST NAME: KORIAL

DATE/TIME OF TEST: 11/27/2017 05:24:27

Diagnostics Check	OK	05:17
Air Blank	0.000	05:18
Control Test	0.081	05:18
Air Blank	0.000	05:19
Subject Sample # 1	0.138	05:19
Breath Volume	2.585	05:19
Air Blank	0.000	05:20
Air Blank	0.000	05:22
Subject Sample # 2	0.143	05:22
Breath Volume	3.406	05:22
Air Blank	0.000	05:23
Control Test	0.080	05:23
Air Blank	0.000	05:24
Diagnostics Check	OK	05:24

Data Download Date/Time: 12/11/2017 15:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001110

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRUZ,RUSSELL,W
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: CRUZ,RUSSELL
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 10/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 620051
DRY GAS EXPIRY DATE: 11/05/2019
VIOLATION CODE: DUI
ARREST TIME: 12:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:06

SUBJECT'S LAST NAME: ROBLERO ROBLERO

DATE/TIME OF TEST: 11/03/2017 01:32:15

Diagnostics Check	OK	01:28	
Air Blank	0.000	01:28	
Control Test	0.081	01:29	
Air Blank	0.000	01:29	
Subject Sample	# 1	01:30	Subject Test Refused
Breath Volume	0.000	01:30	
Air Blank	0.000	01:31	
Control Test	0.079	01:31	
Air Blank	0.000	01:32	
Diagnostics Check	OK	01:32	

Data Download Date/Time: 11/29/2017 09:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001110

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASTILLO,JOSUET,
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: CASTILLO,JOSUET
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 10/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 620051
DRY GAS EXPIRY DATE: 11/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:18

SUBJECT'S LAST NAME: RUSSELL-SOUTHERLAND

DATE/TIME OF TEST: 11/04/2017 02:47:26

Diagnostics Check	OK	02:40
Air Blank	0.000	02:40
Control Test	0.081	02:40
Air Blank	0.000	02:41
Subject Sample # 1	0.167	02:42
Breath Volume	1.390	02:42
Air Blank	0.000	02:42
Air Blank	0.000	02:44
Subject Sample # 2	0.167	02:45
Breath Volume	1.167	02:45
Air Blank	0.000	02:46
Control Test	0.081	02:46
Air Blank	0.000	02:47
Diagnostics Check	OK	02:47

Data Download Date/Time: 11/29/2017 09:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001110

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROCA,LUIS,C
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: ROCA,LUIS
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 10/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 620051
DRY GAS EXPIRY DATE: 11/05/2019
VIOLATION CODE: DUI
ARREST TIME: 07:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:48

SUBJECT'S LAST NAME: VALLE CASTRO

DATE/TIME OF TEST: 11/04/2017 08:38:00

Diagnostics Check	OK	08:31
Air Blank	0.000	08:32
Control Test	0.080	08:32
Air Blank	0.000	08:33
Subject Sample # 1	0.135	08:33
Breath Volume	1.562	08:33
Air Blank	0.000	08:34
Air Blank	0.000	08:35
Subject Sample # 2	0.133	08:36
Breath Volume	1.996	08:36
Air Blank	0.000	08:37
Control Test	0.081	08:37
Air Blank	0.000	08:37
Diagnostics Check	OK	08:38

Data Download Date/Time: 11/29/2017 09:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001110

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASTILLO,JOSUET,
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: CASTILLO,JOSUET
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 10/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 620051
DRY GAS EXPIRY DATE: 11/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:35

SUBJECT'S LAST NAME: PASCUAL

DATE/TIME OF TEST: 11/05/2017 05:02:38

Diagnostics Check	OK	04:55
Air Blank	0.000	04:56
Control Test	0.081	04:56
Air Blank	0.000	04:56
Subject Sample # 1	0.058	04:58
Breath Volume	3.839	04:58
Air Blank	0.000	04:58
Air Blank	0.000	05:00
Subject Sample # 2	0.060	05:01
Breath Volume	3.746	05:01
Air Blank	0.000	05:01
Control Test	0.081	05:02
Air Blank	0.000	05:02
Diagnostics Check	OK	05:02

Data Download Date/Time: 11/29/2017 09:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001110

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LA ROSA CASTRO, DAVID,
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: LA ROSA CASTRO, DAVID
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 10/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 620051
DRY GAS EXPIRY DATE: 11/05/2019
VIOLATION CODE: DUI
ARREST TIME: 05:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:10

SUBJECT'S LAST NAME: MCKAY

DATE/TIME OF TEST: 11/12/2017 06:42:56

Diagnostics Check	OK	06:31	
Air Blank	0.000	06:31	
Control Test	0.081	06:32	
Air Blank	0.000	06:32	
Subject Sample	# 1 0.000	06:35	Volume Not Met
Breath Volume	0.421	06:35	
Air Blank	0.000	06:36	
Air Blank	0.000	06:38	
Subject Sample	# 2 0.000	06:41	Volume Not Met
Breath Volume	0.269	06:41	
Air Blank	0.000	06:42	
Control Test	0.082	06:42	
Air Blank	0.000	06:42	
Diagnostics Check	OK	06:42	

Data Download Date/Time: 11/29/2017 09:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001110

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROCA,LUIS,C
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: ROCA,LUIS
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 10/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 620051
DRY GAS EXPIRY DATE: 11/05/2019
VIOLATION CODE: DUI
ARREST TIME: 08:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:20

SUBJECT'S LAST NAME: MARIACA ARIAS

DATE/TIME OF TEST: 11/13/2017 11:29:31

Diagnostics Check	OK	11:26	
Air Blank	0.000	11:27	
Control Test	0.081	11:27	
Air Blank	0.000	11:27	
Subject Sample	# 1	11:28	Subject Test Refused
Breath Volume	0.000	11:28	
Air Blank	0.000	11:28	
Control Test	0.082	11:28	
Air Blank	0.000	11:29	
Diagnostics Check	OK	11:29	

Data Download Date/Time: 11/29/2017 09:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001110

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIAZ,MARCOS,E
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: DIAZ,MARCOS
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 10/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 620051
DRY GAS EXPIRY DATE: 11/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:02

SUBJECT'S LAST NAME: DEGEN

DATE/TIME OF TEST: 11/18/2017 02:33:37

Diagnostics Check	OK	02:30	
Air Blank	0.000	02:31	
Control Test	0.081	02:31	
Air Blank	0.000	02:31	
Subject Sample	# 1	02:32	Subject Test Refused
Breath Volume	0.000	02:32	
Air Blank	0.000	02:32	
Control Test	0.081	02:33	
Air Blank	0.000	02:33	
Diagnostics Check	OK	02:33	

Data Download Date/Time: 11/29/2017 09:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001110

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROCA,LUIS,C
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: ROCA,LUIS
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 10/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 620051
DRY GAS EXPIRY DATE: 11/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:15

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 11/18/2017 04:49:16

Diagnostics Check	OK	04:39	
Air Blank	0.000	04:39	
Control Test	0.081	04:40	
Air Blank	0.000	04:40	
Subject Sample	# 1 0.232	04:42	
Breath Volume	2.011	04:42	
Air Blank	0.000	04:42	
Air Blank	0.000	04:44	
Subject Sample	# 2	04:47	Subject Test Refused
Breath Volume	0.113	04:47	
Air Blank	0.000	04:48	
Control Test	0.080	04:48	
Air Blank	0.000	04:49	
Diagnostics Check	OK	04:49	

Data Download Date/Time: 11/29/2017 09:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001111

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SISK,CARLOS,M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: SISK,CARLOS
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 11/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 712031
DRY GAS EXPIRY DATE: 02/21/2020
VIOLATION CODE: DUI
ARREST TIME: 05:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:15

SUBJECT'S LAST NAME: BARRON

DATE/TIME OF TEST: 11/22/2017 06:40:30

Diagnostics Check	OK	06:36	
Air Blank	0.000	06:37	
Control Test	0.080	06:37	
Air Blank	0.000	06:38	
Subject Sample	# 1	06:39	Subject Test Refused
Breath Volume	0.000	06:39	
Air Blank	0.000	06:39	
Control Test	0.081	06:39	
Air Blank	0.000	06:40	
Diagnostics Check	OK	06:40	

Data Download Date/Time: 12/01/2017 22:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001112

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWARD, DOMINIQUE, L
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: HOWARD, DOMINIQUE
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 10/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 21:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: MCNANNA

DATE/TIME OF TEST: 11/02/2017 22:51:50

Diagnostics Check	OK	22:45
Air Blank	0.000	22:45
Control Test	0.082	22:45
Air Blank	0.000	22:46
Subject Sample # 1	0.029	22:47
Breath Volume	2.496	22:47
Air Blank	0.000	22:48
Air Blank	0.000	22:49
Subject Sample # 2	0.028	22:50
Breath Volume	1.785	22:50
Air Blank	0.000	22:50
Control Test	0.082	22:51
Air Blank	0.000	22:51
Diagnostics Check	OK	22:51

Data Download Date/Time: 11/28/2017 04:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001112

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOLFF,ROSE,M
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: WOLFF,ROSE
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 10/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 14:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:10

SUBJECT'S LAST NAME: RUBAS

DATE/TIME OF TEST: 11/08/2017 15:38:17

Diagnostics Check	OK	15:35	
Air Blank	0.000	15:35	
Control Test	0.081	15:35	
Air Blank	0.000	15:36	
Subject Sample	# 1	15:36	Subject Test Refused
Breath Volume	0.000	15:36	
Air Blank	0.000	15:37	
Control Test	0.081	15:37	
Air Blank	0.000	15:38	
Diagnostics Check	OK	15:38	

Data Download Date/Time: 11/28/2017 04:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001112

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VALDES,DAVID,S
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: VALDES,DAVID
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 10/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 14:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:54

SUBJECT'S LAST NAME: ALTUS

DATE/TIME OF TEST: 11/10/2017 15:31:53

Diagnostics Check	OK	15:25
Air Blank	0.000	15:26
Control Test	0.081	15:26
Air Blank	0.000	15:26
Subject Sample # 1	0.238	15:27
Breath Volume	3.160	15:27
Air Blank	0.000	15:28
Air Blank	0.000	15:29
Subject Sample # 2	0.239	15:30
Breath Volume	3.605	15:30
Air Blank	0.000	15:30
Control Test	0.081	15:31
Air Blank	0.000	15:31
Diagnostics Check	OK	15:31

Data Download Date/Time: 11/28/2017 04:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001112

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWARD, DOMINIQUE, L
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: HOWARD, DOMINIQUE
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 10/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 17:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:09

SUBJECT'S LAST NAME: LENZ

DATE/TIME OF TEST: 11/26/2017 18:40:36

Diagnostics Check	OK	18:33
Air Blank	0.000	18:34
Control Test	0.081	18:34
Air Blank	0.000	18:35
Subject Sample # 1	0.202	18:35
Breath Volume	2.648	18:35
Air Blank	0.000	18:36
Air Blank	0.000	18:38
Subject Sample # 2	0.198	18:39
Breath Volume	2.539	18:39
Air Blank	0.000	18:39
Control Test	0.082	18:40
Air Blank	0.000	18:40
Diagnostics Check	OK	18:40

Data Download Date/Time: 11/28/2017 04:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001112

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOLFF,ROSE,M
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: MCNEELY,STEVE
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 11/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 19:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:30

SUBJECT'S LAST NAME: BLUE

DATE/TIME OF TEST: 11/28/2017 21:14:50

Diagnostics Check	OK	21:03	
Air Blank	0.000	21:03	
Control Test	0.082	21:04	
Air Blank	0.000	21:04	
Subject Sample	# 1	21:07	No Sample Provided
Breath Volume	0.000	21:07	
Air Blank	0.000	21:08	
Air Blank	0.000	21:10	
Subject Sample	# 2	21:13	No Sample Provided
Breath Volume	0.000	21:13	
Air Blank	0.000	21:13	
Control Test	0.082	21:14	
Air Blank	0.000	21:14	
Diagnostics Check	OK	21:14	

Data Download Date/Time: 12/31/2017 04:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001113

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,ALAIN,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: HERNANDEZ,ALAIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 10/03/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 03:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:03

SUBJECT'S LAST NAME: DUQUE

DATE/TIME OF TEST: 11/04/2017 03:30:53

Diagnostics Check	OK	03:24
Air Blank	0.000	03:24
Control Test	0.083	03:24
Air Blank	0.000	03:25
Subject Sample # 1	0.199	03:26
Breath Volume	1.316	03:26
Air Blank	0.000	03:27
Air Blank	0.000	03:28
Subject Sample # 2	0.197	03:29
Breath Volume	1.878	03:29
Air Blank	0.000	03:29
Control Test	0.082	03:30
Air Blank	0.000	03:30
Diagnostics Check	OK	03:30

Data Download Date/Time: 11/12/2017 12:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001114

INSTRUMENT REGISTERED TO: DELAND P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: FLORYANCE,JASON,R

BREATH TEST OPERATOR AGENCY: Deland PD

ARRESTING OFFICER: LIZABET,EDUARD

ARRESTING OFFICER AGENCY: Deland PD

LAST AGENCY INSPECTION: 11/26/2017

Date or Time Changed: No

DRY GAS STANDARD LOT #: AG616503

DRY GAS EXPIRY DATE: 06/13/2018

VIOLATION CODE: DUI

ARREST TIME: 01:52

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:14

SUBJECT'S LAST NAME: YAT RODRIGUEZ

DATE/TIME OF TEST: 11/27/2017 02:44:39

Diagnostics Check	OK	02:38
Air Blank	0.000	02:38
Control Test	0.078	02:38
Air Blank	0.000	02:39
Subject Sample # 1	0.193	02:40
Breath Volume	1.906	02:40
Air Blank	0.000	02:40
Air Blank	0.000	02:42
Subject Sample # 2	0.181	02:43
Breath Volume	1.980	02:43
Air Blank	0.000	02:43
Control Test	0.077	02:44
Air Blank	0.000	02:44
Diagnostics Check	OK	02:44

Data Download Date/Time: 12/30/2017 14:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001115

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEAN,DONALD,E
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: DEAN,DONALD
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 10/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A6
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 16:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:20

SUBJECT'S LAST NAME: COSLEY

DATE/TIME OF TEST: 11/03/2017 21:40:31

Diagnostics Check	OK	21:34
Air Blank	0.000	21:35
Control Test	0.078	21:35
Air Blank	0.000	21:35
Subject Sample # 1	0.092	21:36
Breath Volume	3.328	21:36
Air Blank	0.000	21:36
Air Blank	0.000	21:38
Subject Sample # 2	0.090	21:39
Breath Volume	3.238	21:39
Air Blank	0.000	21:39
Control Test	0.078	21:39
Air Blank	0.000	21:40
Diagnostics Check	OK	21:40

Data Download Date/Time: 11/10/2017 10:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001115

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NORTHWAY,GEORGIA,K
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: NORTHWAY,GEORGIA
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 10/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A6
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:51

SUBJECT'S LAST NAME: DIXON

DATE/TIME OF TEST: 11/07/2017 23:20:32

Diagnostics Check	OK	23:13
Air Blank	0.000	23:14
Control Test	0.079	23:14
Air Blank	0.000	23:15
Subject Sample # 1	0.147	23:16
Breath Volume	2.824	23:16
Air Blank	0.000	23:16
Air Blank	0.000	23:18
Subject Sample # 2	0.143	23:19
Breath Volume	2.917	23:19
Air Blank	0.000	23:19
Control Test	0.078	23:19
Air Blank	0.000	23:20
Diagnostics Check	OK	23:20

Data Download Date/Time: 11/10/2017 10:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001115

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NORTHWAY,GEORGIA,K
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: NORTHWAY,GEORGIA
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 10/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A6
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:23

SUBJECT'S LAST NAME: CARROLL

DATE/TIME OF TEST: 11/08/2017 06:55:49

Diagnostics Check	OK	06:49
Air Blank	0.000	06:49
Control Test	0.079	06:50
Air Blank	0.000	06:50
Subject Sample # 1	0.088	06:51
Breath Volume	2.792	06:51
Air Blank	0.000	06:51
Air Blank	0.000	06:53
Subject Sample # 2	0.087	06:54
Breath Volume	3.089	06:54
Air Blank	0.000	06:54
Control Test	0.078	06:55
Air Blank	0.000	06:55
Diagnostics Check	OK	06:55

Data Download Date/Time: 11/10/2017 10:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001115

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KAUFFMAN, MITCHELL, P
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: KAUFFMAN, MITCHELL
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 10/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A6
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:12

SUBJECT'S LAST NAME: STRONG

DATE/TIME OF TEST: 11/09/2017 22:39:23

Diagnostics Check	OK	22:33
Air Blank	0.000	22:33
Control Test	0.076	22:34
Air Blank	0.000	22:34
Subject Sample # 1	0.178	22:34
Breath Volume	2.882	22:34
Air Blank	0.000	22:35
Air Blank	0.000	22:37
Subject Sample # 2	0.179	22:37
Breath Volume	2.253	22:37
Air Blank	0.000	22:38
Control Test	0.077	22:38
Air Blank	0.000	22:39
Diagnostics Check	OK	22:39

Data Download Date/Time: 11/10/2017 10:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001118

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TARTER,JEFFREY,A
BREATH TEST OPERATOR AGENCY: Jackson County SO
ARRESTING OFFICER: TARTER,JEFFREY
ARRESTING OFFICER AGENCY: Jackson County SO
LAST AGENCY INSPECTION: 11/03/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 14:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:47

SUBJECT'S LAST NAME: MOUNT

DATE/TIME OF TEST: 11/08/2017 16:13:38

Diagnostics Check	OK	16:07
Air Blank	0.000	16:07
Control Test	0.081	16:07
Air Blank	0.000	16:08
Subject Sample # 1	0.094	16:08
Breath Volume	1.289	16:08
Air Blank	0.000	16:09
Air Blank	0.000	16:11
Subject Sample # 2	0.099	16:12
Breath Volume	2.269	16:12
Air Blank	0.000	16:12
Control Test	0.080	16:13
Air Blank	0.000	16:13
Diagnostics Check	OK	16:13

Data Download Date/Time: 12/05/2017 10:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001118

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COWAN,RICHARD,T
BREATH TEST OPERATOR AGENCY: Jackson County SO
ARRESTING OFFICER: PARK,DAVID
ARRESTING OFFICER AGENCY: Graceville PD
LAST AGENCY INSPECTION: 11/03/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:46

SUBJECT'S LAST NAME: HARRISON

DATE/TIME OF TEST: 11/13/2017 00:16:44

Diagnostics Check	OK	00:06
Air Blank	0.000	00:07
Control Test	0.079	00:07
Air Blank	0.000	00:07
Subject Sample # 1	0.198	00:08
Breath Volume	1.863	00:08
Air Blank	0.000	00:08
Air Blank	0.000	00:10
Subject Sample # 2	0.174	00:12
Breath Volume	1.625	00:12
Air Blank	0.000	00:13
Air Blank	0.000	00:14
Subject Sample # 3	0.188	00:15
Breath Volume	2.117	00:15
Air Blank	0.000	00:15
Control Test	0.079	00:16
Air Blank	0.000	00:16
Diagnostics Check	OK	00:16

Data Download Date/Time: 12/05/2017 10:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001118

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COWAN,RICHARD,T
BREATH TEST OPERATOR AGENCY: Jackson County SO
ARRESTING OFFICER: LEE,TREVOR
ARRESTING OFFICER AGENCY: Jackson County SO
LAST AGENCY INSPECTION: 11/03/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:01

SUBJECT'S LAST NAME: GERALD

DATE/TIME OF TEST: 11/16/2017 01:30:09

Diagnostics Check	OK	01:22
Air Blank	0.000	01:22
Control Test	0.080	01:23
Air Blank	0.000	01:23
Subject Sample # 1	0.170	01:25
Breath Volume	1.855	01:25
Air Blank	0.000	01:25
Air Blank	0.000	01:27
Subject Sample # 2	0.170	01:28
Breath Volume	2.300	01:28
Air Blank	0.000	01:29
Control Test	0.077	01:29
Air Blank	0.000	01:30
Diagnostics Check	OK	01:30

Data Download Date/Time: 12/05/2017 10:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001118

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARGE,SUSAN,M
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: BARGE,SUSAN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 11/03/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:07

SUBJECT'S LAST NAME: GOELZ

DATE/TIME OF TEST: 11/23/2017 21:36:52

Diagnostics Check	OK	21:28	
Air Blank	0.000	21:28	
Control Test	0.081	21:29	
Air Blank	0.000	21:29	
Subject Sample	# 1	21:32	No Sample Provided
Breath Volume	0.000	21:32	
Air Blank	0.000	21:33	
Air Blank	0.000	21:35	
Subject Sample	# 2	21:35	Subject Test Refused
Breath Volume	0.000	21:35	
Air Blank	0.000	21:35	
Control Test	0.082	21:36	
Air Blank	0.000	21:36	
Diagnostics Check	OK	21:36	

Data Download Date/Time: 12/05/2017 10:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001119

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZIGLER,RICKIE,R
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: ZIGLER,RICKIE
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 10/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 437738
DRY GAS EXPIRY DATE: 06/08/2019
VIOLATION CODE: DUI
ARREST TIME: 05:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:57

SUBJECT'S LAST NAME: STONESIFER

DATE/TIME OF TEST: 11/05/2017 06:25:46

Diagnostics Check	OK	06:19
Air Blank	0.000	06:20
Control Test	0.080	06:20
Air Blank	0.000	06:20
Subject Sample # 1	0.130	06:21
Breath Volume	2.332	06:21
Air Blank	0.000	06:21
Air Blank	0.000	06:23
Subject Sample # 2	0.128	06:24
Breath Volume	1.550	06:24
Air Blank	0.000	06:24
Control Test	0.080	06:25
Air Blank	0.000	06:25
Diagnostics Check	OK	06:25

Data Download Date/Time: 11/28/2017 18:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001119

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZIGLER,RICKIE,R
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: RODRIGUEZ,DAILYN
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 10/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 437738
DRY GAS EXPIRY DATE: 06/08/2019
VIOLATION CODE: DUI
ARREST TIME: 09:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: BEECH

DATE/TIME OF TEST: 11/05/2017 23:08:21

Diagnostics Check	OK	23:00
Air Blank	0.000	23:00
Control Test	0.081	23:01
Air Blank	0.000	23:01
Subject Sample # 1	0.192	23:03
Breath Volume	1.527	23:03
Air Blank	0.000	23:04
Air Blank	0.000	23:06
Subject Sample # 2	0.197	23:06
Breath Volume	1.980	23:06
Air Blank	0.000	23:07
Control Test	0.080	23:07
Air Blank	0.000	23:08
Diagnostics Check	OK	23:08

Data Download Date/Time: 11/28/2017 18:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001119

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZIGLER,RICKIE,R
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: ZIGLER,RICKIE
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 10/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 437738
DRY GAS EXPIRY DATE: 06/08/2019
VIOLATION CODE: DUI
ARREST TIME: 05:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:40

SUBJECT'S LAST NAME: GOLDSCHMIDT

DATE/TIME OF TEST: 11/19/2017 06:14:27

Diagnostics Check	OK	06:07
Air Blank	0.000	06:07
Control Test	0.081	06:08
Air Blank	0.000	06:08
Subject Sample # 1	0.160	06:09
Breath Volume	1.378	06:09
Air Blank	0.000	06:10
Air Blank	0.000	06:12
Subject Sample # 2	0.158	06:12
Breath Volume	1.390	06:12
Air Blank	0.000	06:13
Control Test	0.081	06:13
Air Blank	0.000	06:14
Diagnostics Check	OK	06:14

Data Download Date/Time: 11/28/2017 18:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001119

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUTLER,TIMOTHY,M
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: BUTLER,TIMOTHY
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 10/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 437738
DRY GAS EXPIRY DATE: 06/08/2019
VIOLATION CODE: DUI
ARREST TIME: 18:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:43

SUBJECT'S LAST NAME: LONDON

DATE/TIME OF TEST: 11/23/2017 19:21:53

Diagnostics Check	OK	19:08	
Air Blank	0.000	19:08	
Control Test	0.079	19:08	
Air Blank	0.000	19:09	
Subject Sample	# 1	19:12	No Sample Provided
Breath Volume	0.000	19:12	
Air Blank	0.000	19:13	
Air Blank	0.000	19:15	
Subject Sample	# 2	0.205	19:16
Breath Volume	2.292	19:16	
Air Blank	0.000	19:17	
Air Blank	0.000	19:18	
Subject Sample	# 3	0.200	19:20
Breath Volume	1.882	19:20	
Air Blank	0.000	19:20	
Control Test	0.080	19:21	
Air Blank	0.000	19:21	
Diagnostics Check	OK	19:21	

Data Download Date/Time: 11/28/2017 18:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001119

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NUNEZ,ANDREW,R
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: ZIGLER,RICKIE
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 10/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 437738
DRY GAS EXPIRY DATE: 06/08/2019
VIOLATION CODE: DUI
ARREST TIME: 23:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:56

SUBJECT'S LAST NAME: HEISS

DATE/TIME OF TEST: 11/27/2017 00:23:20

Diagnostics Check	OK	00:17
Air Blank	0.000	00:17
Control Test	0.079	00:17
Air Blank	0.000	00:18
Subject Sample # 1	0.254	00:18
Breath Volume	3.343	00:18
Air Blank	0.000	00:19
Air Blank	0.000	00:21
Subject Sample # 2	0.254	00:21
Breath Volume	2.757	00:21
Air Blank	0.000	00:22
Control Test	0.080	00:22
Air Blank	0.000	00:23
Diagnostics Check	OK	00:23

Data Download Date/Time: 11/28/2017 18:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001120

INSTRUMENT REGISTERED TO: FHP TROOP K

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HAUGHIE,DARIN,A

BREATH TEST OPERATOR AGENCY: FHP Troop K

ARRESTING OFFICER: HAUGHIE,DARIN

ARRESTING OFFICER AGENCY: FHP Troop K

LAST AGENCY INSPECTION: 10/26/2017

Date or Time Changed: No

DRY GAS STANDARD LOT #: 366849

DRY GAS EXPIRY DATE: 03/30/2019

VIOLATION CODE: DUI

ARREST TIME: 07:06

GENDER: Female

OBSERVATION PERIOD BEGAN: 07:05

SUBJECT'S LAST NAME: SNOW

DATE/TIME OF TEST: 11/16/2017 08:50:30

Diagnostics Check	OK	08:47	
Air Blank	0.000	08:47	
Control Test	0.079	08:47	
Air Blank	0.000	08:48	
Subject Sample	# 1	08:49	Subject Test Refused
Breath Volume	0.000	08:49	
Air Blank	0.000	08:49	
Control Test	0.079	08:49	
Air Blank	0.000	08:50	
Diagnostics Check	OK	08:50	

Data Download Date/Time: 11/29/2017 08:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001129

INSTRUMENT REGISTERED TO: PUTNAM COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRENCH, CALEB, L
BREATH TEST OPERATOR AGENCY: Putnam County SO
ARRESTING OFFICER: APPLEBY, THOMAS
ARRESTING OFFICER AGENCY: Putnam County SO
LAST AGENCY INSPECTION: 10/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:43

SUBJECT'S LAST NAME: HOLMAN

DATE/TIME OF TEST: 11/02/2017 02:19:46

Diagnostics Check	OK	02:16	
Air Blank	0.000	02:17	
Control Test	0.080	02:17	
Air Blank	0.000	02:18	
Subject Sample	# 1	02:18	Subject Test Refused
Breath Volume	0.000	02:18	
Air Blank	0.000	02:18	
Control Test	0.080	02:19	
Air Blank	0.000	02:19	
Diagnostics Check	OK	02:19	

Data Download Date/Time: 11/30/2017 23:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001129

INSTRUMENT REGISTERED TO: PUTNAM COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS, TIFFANY, R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: WILLIAMS, TIFFANY
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 10/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: KRIZO

DATE/TIME OF TEST: 11/15/2017 05:02:08

Diagnostics Check	OK	04:55
Air Blank	0.000	04:56
Control Test	0.080	04:56
Air Blank	0.000	04:57
Subject Sample # 1	0.051	04:57
Breath Volume	1.761	04:57
Air Blank	0.000	04:58
Air Blank	0.000	05:00
Subject Sample # 2	0.049	05:00
Breath Volume	1.371	05:00
Air Blank	0.000	05:01
Control Test	0.079	05:01
Air Blank	0.000	05:02
Diagnostics Check	OK	05:02

Data Download Date/Time: 11/30/2017 23:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001129

INSTRUMENT REGISTERED TO: PUTNAM COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS, TIFFANY, R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: WILLIAMS, TIFFANY
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 10/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 11/18/2017 02:04:04

Diagnostics Check	OK	01:57
Air Blank	0.000	01:58
Control Test	0.080	01:58
Air Blank	0.000	01:59
Subject Sample # 1	0.128	01:59
Breath Volume	2.234	01:59
Air Blank	0.000	02:00
Air Blank	0.000	02:02
Subject Sample # 2	0.121	02:02
Breath Volume	1.562	02:02
Air Blank	0.000	02:03
Control Test	0.079	02:03
Air Blank	0.000	02:03
Diagnostics Check	OK	02:04

Data Download Date/Time: 11/30/2017 23:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001129

INSTRUMENT REGISTERED TO: PUTNAM COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRENCH, CALEB, L
BREATH TEST OPERATOR AGENCY: Putnam County SO
ARRESTING OFFICER: CLOUGH, WALTER
ARRESTING OFFICER AGENCY: Putnam County SO
LAST AGENCY INSPECTION: 10/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:31

SUBJECT'S LAST NAME: STUCKER

DATE/TIME OF TEST: 11/30/2017 00:36:54

Diagnostics Check	OK	00:30
Air Blank	0.000	00:30
Control Test	0.080	00:31
Air Blank	0.000	00:31
Subject Sample # 1	0.177	00:32
Breath Volume	1.800	00:32
Air Blank	0.000	00:32
Air Blank	0.000	00:34
Subject Sample # 2	0.171	00:35
Breath Volume	1.304	00:35
Air Blank	0.000	00:36
Control Test	0.080	00:36
Air Blank	0.000	00:36
Diagnostics Check	OK	00:36

Data Download Date/Time: 11/30/2017 23:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001130

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PECK,KEITH,D
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: KING,BRANDON
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 11/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 17:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:25

SUBJECT'S LAST NAME: BALFE

DATE/TIME OF TEST: 11/10/2017 18:13:52

Diagnostics Check	OK	18:05	
Air Blank	0.000	18:05	
Control Test	0.075	18:06	
Air Blank	0.000	18:06	
Subject Sample	# 1 0.281	18:08	
Breath Volume	1.179	18:08	
Air Blank	0.000	18:08	
Air Blank	0.000	18:10	
Subject Sample	# 2 0.263	18:12	
Breath Volume	1.136	18:12	
Air Blank	0.000	18:12	
Control Test	0.073	18:13	Control Outside Tolerance
Air Blank	0.000	18:13	

Data Download Date/Time: 12/28/2017 11:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001130

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PECK,KEITH,D
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: PECK,KEITH
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 11/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 15:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:25

SUBJECT'S LAST NAME: BOULERICE

DATE/TIME OF TEST: 11/12/2017 15:59:12

Diagnostics Check	OK	15:56	
Air Blank	0.000	15:56	
Control Test	0.078	15:57	
Air Blank	0.000	15:57	
Subject Sample	# 1	15:57	Subject Test Refused
Breath Volume	0.000	15:57	
Air Blank	0.000	15:58	
Control Test	0.077	15:58	
Air Blank	0.000	15:59	
Diagnostics Check	OK	15:59	

Data Download Date/Time: 12/28/2017 11:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001130

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARMINGTON, WILLIAM, G
BREATH TEST OPERATOR AGENCY: Ormond Beach PD
ARRESTING OFFICER: MOODY, SIERRA
ARRESTING OFFICER AGENCY: Ormond Beach PD
LAST AGENCY INSPECTION: 11/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 23:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:04

SUBJECT'S LAST NAME: DONINI

DATE/TIME OF TEST: 11/17/2017 00:32:11

Diagnostics Check	OK	00:29	
Air Blank	0.000	00:29	
Control Test	0.077	00:30	
Air Blank	0.000	00:30	
Subject Sample	# 1	00:30	Subject Test Refused
Breath Volume	0.000	00:30	
Air Blank	0.000	00:31	
Control Test	0.077	00:31	
Air Blank	0.000	00:32	
Diagnostics Check	OK	00:32	

Data Download Date/Time: 12/28/2017 11:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001130

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANSARD,JOSHUA,D
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: HANSARD,JOSHUA
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 11/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 19:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: MEDRANO BALTAZAR

DATE/TIME OF TEST: 11/17/2017 20:40:10

Diagnostics Check	OK	20:32
Air Blank	0.000	20:32
Control Test	0.077	20:33
Air Blank	0.000	20:33
Subject Sample # 1	0.210	20:34
Breath Volume	2.195	20:34
Air Blank	0.000	20:35
Air Blank	0.000	20:37
Subject Sample # 2	0.201	20:38
Breath Volume	1.824	20:38
Air Blank	0.000	20:39
Control Test	0.075	20:39
Air Blank	0.000	20:40
Diagnostics Check	OK	20:40

Data Download Date/Time: 12/28/2017 11:45