

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001132

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CALKINS,BRIAN,M
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: PATSCH,JOSHUA
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 11/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 19:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:09

SUBJECT'S LAST NAME: KROLL

DATE/TIME OF TEST: 12/01/2017 20:35:23

Diagnostics Check	OK	20:29
Air Blank	0.000	20:29
Control Test	0.080	20:29
Air Blank	0.000	20:30
Subject Sample # 1	0.178	20:30
Breath Volume	1.156	20:30
Air Blank	0.000	20:31
Air Blank	0.000	20:33
Subject Sample # 2	0.179	20:33
Breath Volume	1.699	20:33
Air Blank	0.000	20:34
Control Test	0.081	20:34
Air Blank	0.000	20:35
Diagnostics Check	OK	20:35

Data Download Date/Time: 12/28/2017 12:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001132

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANSARD,JOSHUA,D
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: MYERS,MATTHEW
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 12/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 17:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:53

SUBJECT'S LAST NAME: LEE

DATE/TIME OF TEST: 12/28/2017 18:19:34

Diagnostics Check		OK	18:13
Air Blank		0.000	18:13
Control Test		0.078	18:14
Air Blank		0.000	18:14
Subject Sample	# 1	0.350	18:15
Breath Volume		2.277	18:15
Air Blank		0.000	18:15
Air Blank		0.000	18:17
Subject Sample	# 2	0.346	18:18
Breath Volume		2.097	18:18
Air Blank		0.000	18:18
Control Test		0.079	18:18
Air Blank		0.000	18:19
Diagnostics Check		OK	18:19

Data Download Date/Time: 01/16/2018 09:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001132

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,MATTHEW,A
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: MYERS,MATTHEW
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 12/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 22:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:48

SUBJECT'S LAST NAME: VOLO

DATE/TIME OF TEST: 12/30/2017 23:20:52

Diagnostics Check	OK	23:13
Air Blank	0.000	23:14
Control Test	0.080	23:14
Air Blank	0.000	23:14
Subject Sample # 1	0.245	23:15
Breath Volume	2.292	23:15
Air Blank	0.000	23:16
Air Blank	0.000	23:18
Subject Sample # 2	0.244	23:19
Breath Volume	2.238	23:19
Air Blank	0.000	23:19
Control Test	0.080	23:20
Air Blank	0.000	23:20
Diagnostics Check	OK	23:20

Data Download Date/Time: 01/16/2018 09:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,JOHNATHAN,R
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: RADO,RYAN
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 11/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 366849
DRY GAS EXPIRY DATE: 03/30/2019
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: RATTIGAN

DATE/TIME OF TEST: 12/11/2017 00:35:25

Diagnostics Check	OK	00:29
Air Blank	0.000	00:29
Control Test	0.079	00:29
Air Blank	0.000	00:30
Subject Sample # 1	0.117	00:30
Breath Volume	2.937	00:30
Air Blank	0.000	00:31
Air Blank	0.000	00:33
Subject Sample # 2	0.115	00:33
Breath Volume	1.464	00:33
Air Blank	0.000	00:34
Control Test	0.078	00:34
Air Blank	0.000	00:35
Diagnostics Check	OK	00:35

Data Download Date/Time: 12/21/2017 16:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILSON,MARK,E
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: WILSON,MARK
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 11/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 366849
DRY GAS EXPIRY DATE: 03/30/2019
VIOLATION CODE: DUI
ARREST TIME: 12:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:12

SUBJECT'S LAST NAME: DIBENEDETTO

DATE/TIME OF TEST: 12/17/2017 02:02:51

Diagnostics Check		OK	01:49	
Air Blank		0.000	01:50	
Control Test		0.077	01:50	
Air Blank		0.000	01:51	
Subject Sample	# 1	0.158	01:54	Volume Not Met
Breath Volume		0.503	01:54	
Air Blank		0.000	01:55	
Air Blank		0.000	01:56	
Subject Sample	# 2	0.134	01:57	
Breath Volume		1.808	01:57	
Air Blank		0.000	01:58	
Air Blank		0.000	02:00	
Subject Sample	# 3	0.144	02:01	
Breath Volume		2.468	02:01	
Air Blank		0.000	02:01	
Control Test		0.076	02:02	
Air Blank		0.000	02:02	
Diagnostics Check		OK	02:02	

Data Download Date/Time: 12/21/2017 16:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YARBOROUGH,ERIC,L
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: BRIGGS,JASON
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 11/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 366849
DRY GAS EXPIRY DATE: 03/30/2019
VIOLATION CODE: DUI
ARREST TIME: 15:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:20

SUBJECT'S LAST NAME: HATIRAS

DATE/TIME OF TEST: 12/20/2017 16:00:41

Diagnostics Check	OK	15:53
Air Blank	0.000	15:53
Control Test	0.077	15:54
Air Blank	0.000	15:54
Subject Sample # 1	0.301	15:55
Breath Volume	1.507	15:55
Air Blank	0.000	15:56
Air Blank	0.000	15:58
Subject Sample # 2	0.308	15:59
Breath Volume	1.296	15:59
Air Blank	0.000	15:59
Control Test	0.076	16:00
Air Blank	0.000	16:00
Diagnostics Check	OK	16:00

Data Download Date/Time: 12/21/2017 16:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001136

INSTRUMENT REGISTERED TO: DELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MULERO,JOHAN,
BREATH TEST OPERATOR AGENCY: Deland PD
ARRESTING OFFICER: KEE,CHRISTOPHER
ARRESTING OFFICER AGENCY: Deland PD
LAST AGENCY INSPECTION: 12/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 83551
DRY GAS EXPIRY DATE: 10/02/2018
VIOLATION CODE: DUI
ARREST TIME: 21:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: NICHOLS

DATE/TIME OF TEST: 12/30/2017 22:28:47

Diagnostics Check		OK	22:22
Air Blank		0.000	22:23
Control Test		0.079	22:23
Air Blank		0.000	22:23
Subject Sample	# 1	0.235	22:24
Breath Volume		2.515	22:24
Air Blank		0.000	22:25
Air Blank		0.000	22:26
Subject Sample	# 2	0.228	22:27
Breath Volume		2.531	22:27
Air Blank		0.000	22:27
Control Test		0.078	22:28
Air Blank		0.000	22:28
Diagnostics Check		OK	22:28

Data Download Date/Time: 01/30/2018 10:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001136

INSTRUMENT REGISTERED TO: DELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MULERO,JOHAN,
BREATH TEST OPERATOR AGENCY: Deland PD
ARRESTING OFFICER: STICKELS,JASON
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 12/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 83551
DRY GAS EXPIRY DATE: 10/02/2018
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:34

SUBJECT'S LAST NAME: CORTES

DATE/TIME OF TEST: 12/31/2017 02:10:15

Diagnostics Check	OK	02:02
Air Blank	0.000	02:03
Control Test	0.079	02:03
Air Blank	0.000	02:03
Subject Sample # 1	0.223	02:05
Breath Volume	1.527	02:05
Air Blank	0.000	02:06
Air Blank	0.000	02:08
Subject Sample # 2	0.226	02:08
Breath Volume	1.968	02:08
Air Blank	0.000	02:09
Control Test	0.078	02:09
Air Blank	0.000	02:10
Diagnostics Check	OK	02:10

Data Download Date/Time: 01/30/2018 10:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SAPP,JOSHUA
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 11/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 562838
DRY GAS EXPIRY DATE: 09/06/2019
VIOLATION CODE: DUI
ARREST TIME: 21:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: TORRES MARIN

DATE/TIME OF TEST: 12/01/2017 22:47:46

Diagnostics Check	OK	22:41
Air Blank	0.000	22:41
Control Test	0.082	22:42
Air Blank	0.000	22:42
Subject Sample # 1	0.109	22:43
Breath Volume	1.328	22:43
Air Blank	0.000	22:43
Air Blank	0.000	22:45
Subject Sample # 2	0.111	22:46
Breath Volume	2.300	22:46
Air Blank	0.000	22:46
Control Test	0.081	22:47
Air Blank	0.000	22:47
Diagnostics Check	OK	22:47

Data Download Date/Time: 12/13/2017 12:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINA,MARMILLY,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: HENNESSY,JONATHAN
ARRESTING OFFICER AGENCY: Sunrise PD
LAST AGENCY INSPECTION: 11/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 562838
DRY GAS EXPIRY DATE: 09/06/2019
VIOLATION CODE: DUI
ARREST TIME: 08:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:35

SUBJECT'S LAST NAME: GALINDO MORO

DATE/TIME OF TEST: 12/03/2017 09:25:13

Diagnostics Check	OK	09:18
Air Blank	0.000	09:19
Control Test	0.081	09:19
Air Blank	0.000	09:19
Subject Sample # 1	0.171	09:20
Breath Volume	2.097	09:20
Air Blank	0.000	09:20
Air Blank	0.000	09:22
Subject Sample # 2	0.162	09:23
Breath Volume	1.750	09:23
Air Blank	0.000	09:24
Control Test	0.081	09:24
Air Blank	0.000	09:25
Diagnostics Check	OK	09:25

Data Download Date/Time: 12/13/2017 12:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FLINK,ANDREW
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 11/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 562838
DRY GAS EXPIRY DATE: 09/06/2019
VIOLATION CODE: DUI
ARREST TIME: 20:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:44

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 12/03/2017 21:16:37

Diagnostics Check	OK	21:10
Air Blank	0.000	21:11
Control Test	0.081	21:11
Air Blank	0.000	21:11
Subject Sample # 1	0.112	21:12
Breath Volume	2.695	21:12
Air Blank	0.000	21:12
Air Blank	0.000	21:14
Subject Sample # 2	0.115	21:15
Breath Volume	1.878	21:15
Air Blank	0.000	21:15
Control Test	0.081	21:16
Air Blank	0.000	21:16
Diagnostics Check	OK	21:16

Data Download Date/Time: 12/13/2017 12:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ROJAS,RANDY
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 11/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 562838
DRY GAS EXPIRY DATE: 09/06/2019
VIOLATION CODE: DUI
ARREST TIME: 00:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:16

SUBJECT'S LAST NAME: DINNALL

DATE/TIME OF TEST: 12/06/2017 01:43:31

Diagnostics Check	OK	01:36
Air Blank	0.000	01:36
Control Test	0.082	01:37
Air Blank	0.000	01:37
Subject Sample # 1	0.161	01:39
Breath Volume	1.847	01:39
Air Blank	0.000	01:39
Air Blank	0.000	01:41
Subject Sample # 2	0.163	01:42
Breath Volume	1.613	01:42
Air Blank	0.000	01:42
Control Test	0.080	01:42
Air Blank	0.000	01:43
Diagnostics Check	OK	01:43

Data Download Date/Time: 12/13/2017 12:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FATHIAN,JAY
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 11/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 562838
DRY GAS EXPIRY DATE: 09/06/2019
VIOLATION CODE: DUI
ARREST TIME: 03:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: CRUZ

DATE/TIME OF TEST: 12/09/2017 04:32:45

Diagnostics Check	OK	04:31	
Air Blank	0.000	04:31	
Control Test	0.083	04:31	
Air Blank		04:32	RFI Detect
Air Blank	0.000	04:32	

Data Download Date/Time: 12/13/2017 12:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FATHIAN,JAY
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 11/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 562838
DRY GAS EXPIRY DATE: 09/06/2019
VIOLATION CODE: DUI
ARREST TIME: 03:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: CRUZ

DATE/TIME OF TEST: 12/09/2017 04:46:21

Diagnostics Check	OK	04:35	
Air Blank	0.000	04:35	
Control Test	0.082	04:35	
Air Blank	0.000	04:36	
Subject Sample	# 1	04:39	No Sample Provided
Breath Volume	0.000	04:39	
Air Blank	0.000	04:39	
Air Blank	0.000	04:41	
Subject Sample	# 2	04:45	No Sample Provided
Breath Volume	0.000	04:45	
Air Blank	0.000	04:45	
Control Test	0.082	04:45	
Air Blank	0.000	04:46	
Diagnostics Check	OK	04:46	

Data Download Date/Time: 12/13/2017 12:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GONTKO,MICHAEL
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 11/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 562838
DRY GAS EXPIRY DATE: 09/06/2019
VIOLATION CODE: DUI
ARREST TIME: 04:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:10

SUBJECT'S LAST NAME: SCOTT-JOHNSON

DATE/TIME OF TEST: 12/09/2017 05:42:37

Diagnostics Check	OK	05:31	
Air Blank	0.000	05:31	
Control Test	0.081	05:32	
Air Blank	0.000	05:32	
Subject Sample	# 1	05:35	No Sample Provided
Breath Volume	0.000	05:35	
Air Blank	0.000	05:36	
Air Blank	0.000	05:38	
Subject Sample	# 2	05:41	No Sample Provided
Breath Volume	0.000	05:41	
Air Blank	0.000	05:41	
Control Test	0.081	05:42	
Air Blank	0.000	05:42	
Diagnostics Check	OK	05:42	

Data Download Date/Time: 12/13/2017 12:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GONTKO,MICHAEL
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 11/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 562838
DRY GAS EXPIRY DATE: 09/06/2019
VIOLATION CODE: DUI
ARREST TIME: 04:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:32

SUBJECT'S LAST NAME: SCOTT-JOHNSON

DATE/TIME OF TEST: 12/09/2017 06:01:51

Diagnostics Check	OK	05:55
Air Blank	0.000	05:56
Control Test	0.082	05:56
Air Blank	0.000	05:56
Subject Sample # 1	0.150	05:57
Breath Volume	1.578	05:57
Air Blank	0.000	05:57
Air Blank	0.000	05:59
Subject Sample # 2	0.152	06:00
Breath Volume	1.855	06:00
Air Blank	0.000	06:00
Control Test	0.081	06:01
Air Blank	0.000	06:01
Diagnostics Check	OK	06:01

Data Download Date/Time: 12/13/2017 12:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINA,MARMILLY,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: DUARTE,CLAUDIA
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 11/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 562838
DRY GAS EXPIRY DATE: 09/06/2019
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: PUENTES BALCAZAR

DATE/TIME OF TEST: 12/10/2017 02:36:05

Diagnostics Check		OK	02:28
Air Blank		0.000	02:29
Control Test		0.082	02:29
Air Blank		0.000	02:29
Subject Sample	# 1	0.245	02:30
Breath Volume		1.273	02:30
Air Blank		0.000	02:30
Air Blank		0.000	02:32
Subject Sample	# 2	0.245	02:34
Breath Volume		1.769	02:34
Air Blank		0.000	02:35
Control Test		0.083	02:35
Air Blank		0.000	02:36
Diagnostics Check		OK	02:36

Data Download Date/Time: 12/13/2017 12:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: DALTON,RYAN
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 11/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 562838
DRY GAS EXPIRY DATE: 09/06/2019
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:04

SUBJECT'S LAST NAME: RIOSMARTINEZ

DATE/TIME OF TEST: 12/10/2017 03:31:20

Diagnostics Check	OK	03:24
Air Blank	0.000	03:24
Control Test	0.082	03:25
Air Blank	0.000	03:25
Subject Sample # 1	0.188	03:26
Breath Volume	2.234	03:26
Air Blank	0.000	03:26
Air Blank	0.000	03:28
Subject Sample # 2	0.190	03:29
Breath Volume	2.132	03:29
Air Blank	0.000	03:30
Control Test	0.082	03:30
Air Blank	0.000	03:31
Diagnostics Check	OK	03:31

Data Download Date/Time: 12/13/2017 12:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GAITAN,ADRIAN
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 11/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 562838
DRY GAS EXPIRY DATE: 09/06/2019
VIOLATION CODE: DUI
ARREST TIME: 03:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:38

SUBJECT'S LAST NAME: MUSKA

DATE/TIME OF TEST: 12/10/2017 04:08:49

Diagnostics Check	OK	04:02
Air Blank	0.000	04:02
Control Test	0.082	04:03
Air Blank	0.000	04:03
Subject Sample # 1	0.208	04:04
Breath Volume	2.019	04:04
Air Blank	0.000	04:04
Air Blank	0.000	04:06
Subject Sample # 2	0.201	04:07
Breath Volume	2.390	04:07
Air Blank	0.000	04:07
Control Test	0.082	04:08
Air Blank	0.000	04:08
Diagnostics Check	OK	04:08

Data Download Date/Time: 12/13/2017 12:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: LUNA,KRISTINA
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 12/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 562838
DRY GAS EXPIRY DATE: 09/06/2019
VIOLATION CODE: DUI
ARREST TIME: 02:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: MALIANOS MORAN

DATE/TIME OF TEST: 12/14/2017 03:23:25

Diagnostics Check		OK	03:16
Air Blank		0.000	03:17
Control Test		0.083	03:17
Air Blank		0.000	03:18
Subject Sample	# 1	0.230	03:18
Breath Volume		1.339	03:18
Air Blank		0.000	03:19
Air Blank		0.000	03:20
Subject Sample	# 2	0.226	03:21
Breath Volume		1.351	03:21
Air Blank		0.000	03:22
Control Test		0.082	03:22
Air Blank		0.000	03:23
Diagnostics Check		OK	03:23

Data Download Date/Time: 01/10/2018 12:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: PLUTA,DARIUSZ
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 12/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 562838
DRY GAS EXPIRY DATE: 09/06/2019
VIOLATION CODE: DUI
ARREST TIME: 18:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:08

SUBJECT'S LAST NAME: ROGAL

DATE/TIME OF TEST: 12/14/2017 19:35:18

Diagnostics Check	OK	19:28
Air Blank	0.000	19:29
Control Test	0.082	19:29
Air Blank	0.000	19:29
Subject Sample # 1	0.214	19:30
Breath Volume	2.875	19:30
Air Blank	0.000	19:31
Air Blank	0.000	19:32
Subject Sample # 2	0.208	19:33
Breath Volume	2.707	19:33
Air Blank	0.000	19:34
Control Test	0.082	19:34
Air Blank	0.000	19:35
Diagnostics Check	OK	19:35

Data Download Date/Time: 01/10/2018 12:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SAPP,JOSHUA
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 12/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 562838
DRY GAS EXPIRY DATE: 09/06/2019
VIOLATION CODE: DUI
ARREST TIME: 21:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:09

SUBJECT'S LAST NAME: EDWARDS

DATE/TIME OF TEST: 12/14/2017 22:36:55

Diagnostics Check	OK	22:30
Air Blank	0.000	22:30
Control Test	0.082	22:31
Air Blank	0.000	22:31
Subject Sample # 1	0.217	22:32
Breath Volume	2.382	22:32
Air Blank	0.000	22:33
Air Blank	0.000	22:34
Subject Sample # 2	0.229	22:35
Breath Volume	2.281	22:35
Air Blank	0.000	22:36
Control Test	0.082	22:36
Air Blank	0.000	22:36
Diagnostics Check	OK	22:36

Data Download Date/Time: 01/10/2018 12:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SAPP,JOSHUA
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 12/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 562838
DRY GAS EXPIRY DATE: 09/06/2019
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:28

SUBJECT'S LAST NAME: VICENTE

DATE/TIME OF TEST: 12/15/2017 03:12:21

Diagnostics Check	OK	03:03
Air Blank	0.000	03:04
Control Test	0.081	03:04
Air Blank	0.000	03:04
Subject Sample # 1	0.272	03:07
Breath Volume	2.250	03:07
Air Blank	0.000	03:07
Air Blank	0.000	03:09
Subject Sample # 2	0.265	03:10
Breath Volume	2.195	03:10
Air Blank	0.000	03:11
Control Test	0.080	03:11
Air Blank	0.000	03:12
Diagnostics Check	OK	03:12

Data Download Date/Time: 01/10/2018 12:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: LUNA,KRISTINA
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 12/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 562838
DRY GAS EXPIRY DATE: 09/06/2019
VIOLATION CODE: DUI
ARREST TIME: 02:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: FIGUEROA

DATE/TIME OF TEST: 12/15/2017 04:05:00

Diagnostics Check	OK	03:58
Air Blank	0.000	03:58
Control Test	0.081	03:58
Air Blank	0.000	03:59
Subject Sample # 1	0.179	04:00
Breath Volume	1.972	04:00
Air Blank	0.000	04:00
Air Blank	0.000	04:02
Subject Sample # 2	0.181	04:03
Breath Volume	1.507	04:03
Air Blank	0.000	04:04
Control Test	0.081	04:04
Air Blank	0.000	04:04
Diagnostics Check	OK	04:04

Data Download Date/Time: 01/10/2018 12:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: AMENGOR,JESSICA
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 12/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 562838
DRY GAS EXPIRY DATE: 09/06/2019
VIOLATION CODE: DUI
ARREST TIME: 03:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:36

SUBJECT'S LAST NAME: WALKER

DATE/TIME OF TEST: 12/16/2017 04:07:33

Diagnostics Check		OK	04:00
Air Blank		0.000	04:00
Control Test		0.081	04:01
Air Blank		0.000	04:01
Subject Sample	# 1	0.084	04:02
Breath Volume		1.707	04:02
Air Blank		0.000	04:02
Air Blank		0.000	04:04
Subject Sample	# 2	0.082	04:06
Breath Volume		1.363	04:06
Air Blank		0.000	04:06
Control Test		0.081	04:06
Air Blank		0.000	04:07
Diagnostics Check		OK	04:07

Data Download Date/Time: 01/10/2018 12:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SCOTT-JENKINS,FELICIA
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 12/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 562838
DRY GAS EXPIRY DATE: 09/06/2019
VIOLATION CODE: DUI
ARREST TIME: 07:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:45

SUBJECT'S LAST NAME: SELLAN MACIAS

DATE/TIME OF TEST: 12/16/2017 09:37:17

Diagnostics Check	OK	09:29
Air Blank	0.000	09:30
Control Test	0.082	09:30
Air Blank	0.000	09:30
Subject Sample # 1	0.119	09:32
Breath Volume	1.660	09:32
Air Blank	0.000	09:33
Air Blank	0.000	09:35
Subject Sample # 2	0.118	09:35
Breath Volume	1.812	09:35
Air Blank	0.000	09:36
Control Test	0.081	09:36
Air Blank	0.000	09:37
Diagnostics Check	OK	09:37

Data Download Date/Time: 01/10/2018 12:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINA,MARMILLY,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: LOUIS,WALEX
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 12/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 562838
DRY GAS EXPIRY DATE: 09/06/2019
VIOLATION CODE: DUI
ARREST TIME: 01:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: ROTHHAUS

DATE/TIME OF TEST: 12/18/2017 02:38:31

Diagnostics Check	OK	02:32
Air Blank	0.000	02:32
Control Test	0.082	02:33
Air Blank	0.000	02:33
Subject Sample # 1	0.129	02:34
Breath Volume	1.355	02:34
Air Blank	0.000	02:34
Air Blank	0.000	02:36
Subject Sample # 2	0.123	02:37
Breath Volume	1.160	02:37
Air Blank	0.000	02:37
Control Test	0.082	02:37
Air Blank	0.000	02:38
Diagnostics Check	OK	02:38

Data Download Date/Time: 01/10/2018 12:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SERPHOS,STEVEN
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 12/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 562838
DRY GAS EXPIRY DATE: 09/06/2019
VIOLATION CODE: DUI
ARREST TIME: 19:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:06

SUBJECT'S LAST NAME: FAVORS III

DATE/TIME OF TEST: 12/19/2017 20:36:12

Diagnostics Check	OK	20:26
Air Blank	0.000	20:27
Control Test	0.082	20:27
Air Blank	0.000	20:27
Subject Sample # 1	0.152	20:29
Breath Volume	1.117	20:29
Air Blank	0.000	20:30
Air Blank	0.000	20:32
Subject Sample # 2	0.152	20:34
Breath Volume	1.281	20:34
Air Blank	0.000	20:35
Control Test	0.081	20:35
Air Blank	0.000	20:36
Diagnostics Check	OK	20:36

Data Download Date/Time: 01/10/2018 12:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: THOMPSON,PATRICK
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 12/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 562838
DRY GAS EXPIRY DATE: 09/06/2019
VIOLATION CODE: DUI
ARREST TIME: 18:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:20

SUBJECT'S LAST NAME: CASARIAN

DATE/TIME OF TEST: 12/21/2017 19:58:44

Diagnostics Check	OK	19:52
Air Blank	0.000	19:52
Control Test	0.082	19:53
Air Blank	0.000	19:53
Subject Sample # 1	0.209	19:54
Breath Volume	1.128	19:54
Air Blank	0.000	19:54
Air Blank	0.000	19:56
Subject Sample # 2	0.209	19:57
Breath Volume	1.136	19:57
Air Blank	0.000	19:57
Control Test	0.081	19:58
Air Blank	0.000	19:58
Diagnostics Check	OK	19:58

Data Download Date/Time: 01/10/2018 12:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ROJAS,RICKY
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 12/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 562838
DRY GAS EXPIRY DATE: 09/06/2019
VIOLATION CODE: DUI
ARREST TIME: 01:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:01

SUBJECT'S LAST NAME: TORRES-MALDONADO

DATE/TIME OF TEST: 12/22/2017 02:30:39

Diagnostics Check	OK	02:22
Air Blank	0.000	02:23
Control Test	0.081	02:23
Air Blank	0.000	02:23
Subject Sample # 1	0.138	02:25
Breath Volume	2.507	02:25
Air Blank	0.000	02:26
Air Blank	0.000	02:28
Subject Sample # 2	0.139	02:29
Breath Volume	1.988	02:29
Air Blank	0.000	02:29
Control Test	0.081	02:30
Air Blank	0.000	02:30
Diagnostics Check	OK	02:30

Data Download Date/Time: 01/10/2018 12:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SAPP,JOSHUA
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 12/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 562838
DRY GAS EXPIRY DATE: 09/06/2019
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:28

SUBJECT'S LAST NAME: JUCHELKA

DATE/TIME OF TEST: 12/23/2017 02:55:57

Diagnostics Check	OK	02:49
Air Blank	0.000	02:49
Control Test	0.081	02:49
Air Blank	0.000	02:50
Subject Sample # 1	0.227	02:50
Breath Volume	1.109	02:50
Air Blank	0.000	02:51
Air Blank	0.000	02:53
Subject Sample # 2	0.221	02:54
Breath Volume	2.179	02:54
Air Blank	0.000	02:55
Control Test	0.082	02:55
Air Blank	0.000	02:55
Diagnostics Check	OK	02:55

Data Download Date/Time: 01/10/2018 12:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ALONSO,BERNARDO
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 12/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 562838
DRY GAS EXPIRY DATE: 09/06/2019
VIOLATION CODE: DUI
ARREST TIME: 04:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:28

SUBJECT'S LAST NAME: MARTIN

DATE/TIME OF TEST: 12/23/2017 04:54:27

Diagnostics Check	OK	04:48
Air Blank	0.000	04:49
Control Test	0.082	04:49
Air Blank	0.000	04:49
Subject Sample # 1	0.000	04:50
Breath Volume	2.796	04:50
Air Blank	0.000	04:50
Air Blank	0.000	04:52
Subject Sample # 2	0.000	04:53
Breath Volume	2.539	04:53
Air Blank	0.000	04:53
Control Test	0.081	04:53
Air Blank	0.000	04:54
Diagnostics Check	OK	04:54

Data Download Date/Time: 01/10/2018 12:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINA,MARMILLY,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: LOUIS,WALEX
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 12/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 562838
DRY GAS EXPIRY DATE: 09/06/2019
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: CHERISMA

DATE/TIME OF TEST: 12/25/2017 03:11:09

Diagnostics Check	OK	03:04
Air Blank	0.000	03:04
Control Test	0.082	03:05
Air Blank	0.000	03:05
Subject Sample # 1	0.167	03:06
Breath Volume	1.902	03:06
Air Blank	0.000	03:07
Air Blank	0.000	03:09
Subject Sample # 2	0.170	03:09
Breath Volume	1.152	03:09
Air Blank	0.000	03:10
Control Test	0.081	03:10
Air Blank	0.000	03:11
Diagnostics Check	OK	03:11

Data Download Date/Time: 01/10/2018 12:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINA,MARMILLY,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: KOTT,DOUGLAS
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 12/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 562838
DRY GAS EXPIRY DATE: 09/06/2019
VIOLATION CODE: DUI
ARREST TIME: 23:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: GRABLE

DATE/TIME OF TEST: 12/26/2017 00:57:17

Diagnostics Check		OK	00:43	
Air Blank		0.000	00:43	
Control Test		0.082	00:44	
Air Blank		0.000	00:44	
Subject Sample	# 1	0.162	00:47	Volume Not Met
Breath Volume		0.019	00:47	
Air Blank		0.000	00:48	
Air Blank		0.000	00:50	
Subject Sample	# 2	0.181	00:51	
Breath Volume		1.195	00:51	
Air Blank		0.000	00:52	
Air Blank		0.000	00:54	
Subject Sample	# 3	0.166	00:55	
Breath Volume		1.269	00:55	
Air Blank		0.000	00:56	
Control Test		0.081	00:56	
Air Blank		0.000	00:57	
Diagnostics Check		OK	00:57	

Data Download Date/Time: 01/10/2018 12:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: BUDISH,HARRISON
ARRESTING OFFICER AGENCY: Pembroke Pines PD
LAST AGENCY INSPECTION: 12/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 562838
DRY GAS EXPIRY DATE: 09/06/2019
VIOLATION CODE: DUI
ARREST TIME: 03:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:12

SUBJECT'S LAST NAME: PAJOTTE

DATE/TIME OF TEST: 12/26/2017 04:38:40

Diagnostics Check	OK	04:32
Air Blank	0.000	04:32
Control Test	0.082	04:33
Air Blank	0.000	04:33
Subject Sample # 1	0.079	04:34
Breath Volume	1.289	04:34
Air Blank	0.000	04:34
Air Blank	0.000	04:36
Subject Sample # 2	0.079	04:37
Breath Volume	2.199	04:37
Air Blank	0.000	04:37
Control Test	0.082	04:38
Air Blank	0.000	04:38
Diagnostics Check	OK	04:38

Data Download Date/Time: 01/10/2018 12:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAPP,JOSHUA,R
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ORTIZ,JEFFREY
ARRESTING OFFICER AGENCY: Hollywood PD
LAST AGENCY INSPECTION: 12/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 562838
DRY GAS EXPIRY DATE: 09/06/2019
VIOLATION CODE: DUI
ARREST TIME: 19:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:55

SUBJECT'S LAST NAME: A416780670500

DATE/TIME OF TEST: 12/28/2017 21:33:10

Diagnostics Check	OK	21:25
Air Blank	0.000	21:26
Control Test	0.082	21:26
Air Blank	0.000	21:26
Subject Sample # 1	0.179	21:27
Breath Volume	1.218	21:27
Air Blank	0.000	21:28
Air Blank	0.000	21:30
Subject Sample # 2	0.188	21:31
Breath Volume	1.980	21:31
Air Blank	0.000	21:32
Control Test	0.082	21:32
Air Blank	0.000	21:33
Diagnostics Check	OK	21:33

Data Download Date/Time: 01/10/2018 12:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINA,MARMILLY,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: HENNESSY,JONATHAN
ARRESTING OFFICER AGENCY: Sunrise PD
LAST AGENCY INSPECTION: 12/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 562838
DRY GAS EXPIRY DATE: 09/06/2019
VIOLATION CODE: DUI
ARREST TIME: 06:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:05

SUBJECT'S LAST NAME: VALENCIA

DATE/TIME OF TEST: 12/29/2017 08:53:09

Diagnostics Check		OK	08:41	
Air Blank		0.000	08:42	
Control Test		0.082	08:42	
Air Blank		0.000	08:42	
Subject Sample	# 1	0.118	08:46	Volume Not Met
Breath Volume		0.234	08:46	
Air Blank		0.000	08:46	
Air Blank		0.000	08:48	
Subject Sample	# 2	0.127	08:51	Volume Not Met
Breath Volume		0.636	08:51	
Air Blank		0.000	08:52	
Control Test		0.083	08:52	
Air Blank		0.000	08:53	
Diagnostics Check		OK	08:53	

Data Download Date/Time: 01/10/2018 12:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINA,MARMILLY,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: IRELAN,ANDREW
ARRESTING OFFICER AGENCY: Margate PD
LAST AGENCY INSPECTION: 12/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 562838
DRY GAS EXPIRY DATE: 09/06/2019
VIOLATION CODE: DUI
ARREST TIME: 01:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: MCALPIN

DATE/TIME OF TEST: 12/30/2017 02:56:10

Diagnostics Check	OK	02:44
Air Blank	0.000	02:45
Control Test	0.082	02:45
Air Blank	0.000	02:46
Subject Sample # 1	0.243	02:46
Breath Volume	1.164	02:46
Air Blank	0.000	02:47
Air Blank	0.000	02:49
Subject Sample # 2	0.218	02:51
Breath Volume	1.363	02:51
Air Blank	0.000	02:52
Air Blank	0.000	02:54
Subject Sample # 3	0.220	02:54
Breath Volume	1.304	02:54
Air Blank	0.000	02:55
Control Test	0.081	02:55
Air Blank	0.000	02:56
Diagnostics Check	OK	02:56

Data Download Date/Time: 01/10/2018 12:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINA,MARMILLY,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: TRABUE,THOMAS
ARRESTING OFFICER AGENCY: Pembroke Pines PD
LAST AGENCY INSPECTION: 12/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 562838
DRY GAS EXPIRY DATE: 09/06/2019
VIOLATION CODE: DUI
ARREST TIME: 08:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:30

SUBJECT'S LAST NAME: NIEVES

DATE/TIME OF TEST: 12/30/2017 10:08:45

Diagnostics Check	OK	10:02
Air Blank	0.000	10:02
Control Test	0.082	10:03
Air Blank	0.000	10:03
Subject Sample # 1	0.179	10:04
Breath Volume	2.242	10:04
Air Blank	0.000	10:04
Air Blank	0.000	10:06
Subject Sample # 2	0.182	10:07
Breath Volume	2.199	10:07
Air Blank	0.000	10:07
Control Test	0.082	10:08
Air Blank	0.000	10:08
Diagnostics Check	OK	10:08

Data Download Date/Time: 01/10/2018 12:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINA,MARMILLY,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: HALLER,CARL
ARRESTING OFFICER AGENCY: Sunrise PD
LAST AGENCY INSPECTION: 12/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 562838
DRY GAS EXPIRY DATE: 09/06/2019
VIOLATION CODE: DUI
ARREST TIME: 02:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: SOUSA

DATE/TIME OF TEST: 12/31/2017 03:22:17

Diagnostics Check	OK	03:17	
Air Blank	0.000	03:18	
Control Test	0.081	03:18	
Air Blank	0.000	03:19	
Subject Sample	# 1	03:20	Subject Test Refused
Breath Volume	0.000	03:20	
Air Blank	0.000	03:21	
Control Test	0.082	03:21	
Air Blank	0.000	03:22	
Diagnostics Check	OK	03:22	

Data Download Date/Time: 01/10/2018 12:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINA,MARMILLY,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SAINT AMOUR,JEAN
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 12/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 562838
DRY GAS EXPIRY DATE: 09/06/2019
VIOLATION CODE: DUI
ARREST TIME: 07:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:00

SUBJECT'S LAST NAME: CONSTANT

DATE/TIME OF TEST: 12/31/2017 09:39:02

Diagnostics Check	OK	09:32
Air Blank	0.000	09:32
Control Test	0.082	09:33
Air Blank	0.000	09:33
Subject Sample # 1	0.133	09:34
Breath Volume	1.855	09:34
Air Blank	0.000	09:35
Air Blank	0.000	09:37
Subject Sample # 2	0.140	09:37
Breath Volume	1.734	09:37
Air Blank	0.000	09:38
Control Test	0.082	09:38
Air Blank	0.000	09:38
Diagnostics Check	OK	09:39

Data Download Date/Time: 01/10/2018 12:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINA,MARMILLY,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: DEDEJ,MEDJON
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 12/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 562838
DRY GAS EXPIRY DATE: 09/06/2019
VIOLATION CODE: DUI
ARREST TIME: 09:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:05

SUBJECT'S LAST NAME: ANDRE

DATE/TIME OF TEST: 12/31/2017 10:51:04

Diagnostics Check	OK	10:42	
Air Blank	0.000	10:43	
Control Test	0.082	10:43	
Air Blank	0.000	10:43	
Subject Sample	# 1	10:47	No Sample Provided
Breath Volume	0.000	10:47	
Air Blank	0.000	10:47	
Air Blank	0.000	10:49	
Subject Sample	# 2	10:49	Subject Test Refused
Breath Volume	0.000	10:49	
Air Blank	0.000	10:50	
Control Test	0.082	10:50	
Air Blank	0.000	10:50	
Diagnostics Check	OK	10:51	

Data Download Date/Time: 01/10/2018 12:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POTTER, ADAM, H
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: POTTER, ADAM
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 390314
DRY GAS EXPIRY DATE: 04/19/2019
VIOLATION CODE: DUI
ARREST TIME: 01:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:37

SUBJECT'S LAST NAME: LOPEZ VEGA

DATE/TIME OF TEST: 12/02/2017 02:04:22

Diagnostics Check	OK	01:57
Air Blank	0.000	01:58
Control Test	0.082	01:58
Air Blank	0.000	01:58
Subject Sample # 1	0.140	01:59
Breath Volume	1.433	01:59
Air Blank	0.000	02:00
Air Blank	0.000	02:02
Subject Sample # 2	0.139	02:02
Breath Volume	2.031	02:02
Air Blank	0.000	02:03
Control Test	0.081	02:03
Air Blank	0.000	02:04
Diagnostics Check	OK	02:04

Data Download Date/Time: 12/26/2017 07:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TUCKER, WILLIAM,
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: TUCKER, WILLIAM
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 11/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 390314
DRY GAS EXPIRY DATE: 04/19/2019
VIOLATION CODE: DUI
ARREST TIME: 00:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: KEEL

DATE/TIME OF TEST: 12/03/2017 01:05:58

Diagnostics Check		OK	00:59
Air Blank		0.000	00:59
Control Test		0.081	01:00
Air Blank		0.000	01:00
Subject Sample	# 1	0.229	01:01
Breath Volume		3.695	01:01
Air Blank		0.000	01:01
Air Blank		0.000	01:03
Subject Sample	# 2	0.235	01:04
Breath Volume		3.925	01:04
Air Blank		0.000	01:05
Control Test		0.082	01:05
Air Blank		0.000	01:05
Diagnostics Check		OK	01:05

Data Download Date/Time: 12/26/2017 07:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,FIDEL,
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: TURNER,LATORIAN
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 11/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 390314
DRY GAS EXPIRY DATE: 04/19/2019
VIOLATION CODE: DUI
ARREST TIME: 04:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:56

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 12/10/2017 05:27:52

Diagnostics Check	OK	05:23	
Air Blank	0.000	05:24	
Control Test	0.081	05:24	
Air Blank	0.000	05:24	
Subject Sample	# 1	05:26	Interferent Detect
Breath Volume	1.621	05:26	
Air Blank		05:27	Purge Fail
Air Blank		05:27	Purge Fail

Data Download Date/Time: 12/26/2017 07:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TURNER,LATORIAN,T
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: TURNER,LATORIAN
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 11/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 390314
DRY GAS EXPIRY DATE: 04/19/2019
VIOLATION CODE: DUI
ARREST TIME: 04:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:56

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 12/10/2017 05:34:03

Diagnostics Check	OK	05:30	
Air Blank	0.000	05:30	
Control Test	0.081	05:30	
Air Blank	0.000	05:31	
Subject Sample	# 1	05:32	Subject Test Refused
Breath Volume	0.015	05:32	
Air Blank	0.000	05:33	
Control Test	0.081	05:33	
Air Blank	0.000	05:33	
Diagnostics Check	OK	05:34	

Data Download Date/Time: 12/26/2017 07:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARING,KELLY,A
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: DIAZ,CARLOS
ARRESTING OFFICER AGENCY: Lake Alfred PD
LAST AGENCY INSPECTION: 11/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 390314
DRY GAS EXPIRY DATE: 04/19/2019
VIOLATION CODE: DUI
ARREST TIME: 19:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:20

SUBJECT'S LAST NAME: BAEZ RIVAS

DATE/TIME OF TEST: 12/21/2017 20:57:38

Diagnostics Check	OK	20:54	
Air Blank	0.000	20:54	
Control Test	0.082	20:54	
Air Blank	0.000	20:55	
Subject Sample	# 1	20:55	Interferent Detect
Breath Volume	3.609	20:55	
Air Blank	0.000	20:56	
Control Test	0.081	20:57	Interferent Detect
Air Blank	0.000	20:57	

Data Download Date/Time: 12/26/2017 07:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARING,KELLY,A
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: DIAZ,CARLOS
ARRESTING OFFICER AGENCY: Lake Alfred PD
LAST AGENCY INSPECTION: 11/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 390314
DRY GAS EXPIRY DATE: 04/19/2019
VIOLATION CODE: DUI
ARREST TIME: 19:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:20

SUBJECT'S LAST NAME: BAEZ RIVAS

DATE/TIME OF TEST: 12/21/2017 21:06:49

Diagnostics Check	OK	21:00
Air Blank	0.000	21:01
Control Test	0.081	21:01
Air Blank	0.000	21:01
Subject Sample # 1	0.199	21:02
Breath Volume	3.386	21:02
Air Blank	0.000	21:03
Air Blank	0.000	21:04
Subject Sample # 2	0.208	21:05
Breath Volume	5.363	21:05
Air Blank	0.000	21:05
Control Test	0.079	21:06
Air Blank	0.000	21:06
Diagnostics Check	OK	21:06

Data Download Date/Time: 12/26/2017 07:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TURNER,LATORIAN,T
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: TURNER,LATORIAN
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 11/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 390314
DRY GAS EXPIRY DATE: 04/19/2019
VIOLATION CODE: DUI
ARREST TIME: 16:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:44

SUBJECT'S LAST NAME: HITLEY

DATE/TIME OF TEST: 12/23/2017 17:10:50

Diagnostics Check	OK	17:04
Air Blank	0.000	17:05
Control Test	0.081	17:05
Air Blank	0.000	17:05
Subject Sample # 1	0.216	17:06
Breath Volume	2.789	17:06
Air Blank	0.000	17:07
Air Blank	0.000	17:08
Subject Sample # 2	0.212	17:09
Breath Volume	2.519	17:09
Air Blank	0.000	17:09
Control Test	0.080	17:10
Air Blank	0.000	17:10
Diagnostics Check	OK	17:10

Data Download Date/Time: 12/26/2017 07:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001153

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMIDT,M,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: SMIDT,M
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 11/07/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A6
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: REIS

DATE/TIME OF TEST: 12/01/2017 02:57:36

Diagnostics Check	OK	02:50
Air Blank	0.000	02:50
Control Test	0.080	02:51
Air Blank	0.000	02:51
Subject Sample # 1	0.195	02:52
Breath Volume	1.785	02:52
Air Blank	0.000	02:53
Air Blank	0.000	02:55
Subject Sample # 2	0.196	02:56
Breath Volume	2.320	02:56
Air Blank	0.000	02:56
Control Test	0.080	02:56
Air Blank	0.000	02:57
Diagnostics Check	OK	02:57

Data Download Date/Time: 12/07/2017 11:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001153

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CUTCLIFFE, MARY, P
BREATH TEST OPERATOR AGENCY: Leon County SO
ARRESTING OFFICER: REESE, BRITTNEY
ARRESTING OFFICER AGENCY: Leon County SO
LAST AGENCY INSPECTION: 11/07/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A6
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 17:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:34

SUBJECT'S LAST NAME: DEGEN

DATE/TIME OF TEST: 12/01/2017 19:02:33

Diagnostics Check	OK	18:54
Air Blank	0.000	18:54
Control Test	0.081	18:55
Air Blank	0.000	18:55
Subject Sample # 1	0.070	18:56
Breath Volume	1.246	18:56
Air Blank	0.000	18:57
Air Blank	0.000	18:59
Subject Sample # 2	0.075	19:01
Breath Volume	1.359	19:01
Air Blank	0.000	19:01
Control Test	0.081	19:01
Air Blank	0.000	19:02
Diagnostics Check	OK	19:02

Data Download Date/Time: 12/07/2017 11:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001153

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CROSS, SR.,MICHAEL,S
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: CROSS, SR.,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 11/07/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A6
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: KNIGHT

DATE/TIME OF TEST: 12/03/2017 01:28:39

Diagnostics Check		OK	01:18
Air Blank		0.000	01:18
Control Test		0.078	01:19
Air Blank		0.000	01:19
Subject Sample	# 1	0.277	01:20
Breath Volume		1.562	01:20
Air Blank		0.000	01:21
Air Blank		0.000	01:23
Subject Sample	# 2	0.300	01:23
Breath Volume		1.683	01:23
Air Blank		0.000	01:24
Air Blank		0.000	01:26
Subject Sample	# 3	0.287	01:27
Breath Volume		2.109	01:27
Air Blank		0.000	01:27
Control Test		0.079	01:28
Air Blank		0.000	01:28
Diagnostics Check		OK	01:28

Data Download Date/Time: 12/07/2017 11:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001153

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NORTHWAY,GEORGIA,K
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: NORTHWAY,GEORGIA
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 11/07/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A6
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:57

SUBJECT'S LAST NAME: MAN

DATE/TIME OF TEST: 12/03/2017 02:34:30

Diagnostics Check	OK	02:27
Air Blank	0.000	02:27
Control Test	0.080	02:27
Air Blank	0.000	02:28
Subject Sample # 1	0.131	02:29
Breath Volume	2.468	02:29
Air Blank	0.000	02:30
Air Blank	0.000	02:32
Subject Sample # 2	0.124	02:32
Breath Volume	1.980	02:32
Air Blank	0.000	02:33
Control Test	0.080	02:33
Air Blank	0.000	02:34
Diagnostics Check	OK	02:34

Data Download Date/Time: 12/07/2017 11:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001153

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,PARISE,L
BREATH TEST OPERATOR AGENCY: FL State University
ARRESTING OFFICER: BISHOP,APRIL
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 11/07/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A6
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:37

SUBJECT'S LAST NAME: CRATHER

DATE/TIME OF TEST: 12/04/2017 05:10:08

Diagnostics Check	OK	05:03
Air Blank	0.000	05:04
Control Test	0.081	05:04
Air Blank	0.000	05:05
Subject Sample # 1	0.078	05:05
Breath Volume	2.347	05:05
Air Blank	0.000	05:06
Air Blank	0.000	05:08
Subject Sample # 2	0.074	05:08
Breath Volume	1.785	05:08
Air Blank	0.000	05:09
Control Test	0.080	05:09
Air Blank	0.000	05:10
Diagnostics Check	OK	05:10

Data Download Date/Time: 12/07/2017 11:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001153

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVERETT, MATTHEW, A
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: HARVEY, RHETT
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 11/07/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A6
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: FORD

DATE/TIME OF TEST: 12/07/2017 00:14:16

Diagnostics Check	OK	00:07
Air Blank	0.000	00:07
Control Test	0.081	00:08
Air Blank	0.000	00:08
Subject Sample # 1	0.334	00:09
Breath Volume	1.953	00:09
Air Blank	0.000	00:09
Air Blank	0.000	00:11
Subject Sample # 2	0.329	00:12
Breath Volume	1.937	00:12
Air Blank	0.000	00:13
Control Test	0.081	00:13
Air Blank	0.000	00:14
Diagnostics Check	OK	00:14

Data Download Date/Time: 12/07/2017 11:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001153

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVERETT, MATTHEW, A
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: EDWARDS, JAMES
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 12/07/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A6
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:14

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 12/10/2017 01:45:47

Diagnostics Check		OK	01:39
Air Blank		0.000	01:39
Control Test		0.083	01:40
Air Blank		0.000	01:40
Subject Sample	# 1	0.054	01:41
Breath Volume		1.332	01:41
Air Blank		0.000	01:41
Air Blank		0.000	01:43
Subject Sample	# 2	0.061	01:44
Breath Volume		2.070	01:44
Air Blank		0.000	01:44
Control Test		0.083	01:45
Air Blank		0.000	01:45
Diagnostics Check		OK	01:45

Data Download Date/Time: 01/11/2018 13:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001153

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KELLER, DAVID,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: CONDOR, C
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 12/07/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A6
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 07:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:20

SUBJECT'S LAST NAME: WATSON

DATE/TIME OF TEST: 12/22/2017 09:42:02

Diagnostics Check	OK	09:40	
Air Blank	0.000	09:41	
Control Test	0.081	09:41	
Air Blank		09:41	RFI Detect
Air Blank	0.000	09:42	

Data Download Date/Time: 01/11/2018 13:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001153

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KELLER, DAVID,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: CONDOR, C
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 12/07/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A6
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 07:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:20

SUBJECT'S LAST NAME: WATSON

DATE/TIME OF TEST: 12/22/2017 09:50:35

Diagnostics Check	OK	09:44
Air Blank	0.000	09:44
Control Test	0.081	09:44
Air Blank	0.000	09:45
Subject Sample # 1	0.074	09:45
Breath Volume	1.574	09:45
Air Blank	0.000	09:46
Air Blank	0.000	09:48
Subject Sample # 2	0.071	09:49
Breath Volume	1.718	09:49
Air Blank	0.000	09:49
Control Test	0.081	09:49
Air Blank	0.000	09:50
Diagnostics Check	OK	09:50

Data Download Date/Time: 01/11/2018 13:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001153

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DILLS,JAMES,M
BREATH TEST OPERATOR AGENCY: Leon County SO
ARRESTING OFFICER: DILLS,JAMES
ARRESTING OFFICER AGENCY: Leon County SO
LAST AGENCY INSPECTION: 12/07/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A6
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:21

SUBJECT'S LAST NAME: ALLEN

DATE/TIME OF TEST: 12/27/2017 02:54:37

Diagnostics Check	OK	02:48
Air Blank	0.000	02:49
Control Test	0.083	02:49
Air Blank	0.000	02:49
Subject Sample # 1	0.212	02:50
Breath Volume	2.542	02:50
Air Blank	0.000	02:50
Air Blank	0.000	02:52
Subject Sample # 2	0.210	02:53
Breath Volume	4.589	02:53
Air Blank	0.000	02:53
Control Test	0.081	02:54
Air Blank	0.000	02:54
Diagnostics Check	OK	02:54

Data Download Date/Time: 01/11/2018 13:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001153

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROY,G,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: ROY,G
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 12/07/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A6
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: SUTTON JR

DATE/TIME OF TEST: 12/31/2017 02:50:18

Diagnostics Check	OK	02:44
Air Blank	0.000	02:44
Control Test	0.082	02:44
Air Blank	0.000	02:45
Subject Sample # 1	0.122	02:45
Breath Volume	2.613	02:45
Air Blank	0.000	02:46
Air Blank	0.000	02:48
Subject Sample # 2	0.120	02:48
Breath Volume	3.207	02:48
Air Blank	0.000	02:49
Control Test	0.081	02:49
Air Blank	0.000	02:50
Diagnostics Check	OK	02:50

Data Download Date/Time: 01/11/2018 13:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001154

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANSARD,JOSHUA,D
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: HANSARD,JOSHUA
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 11/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1787697
DRY GAS EXPIRY DATE: 01/08/2018
VIOLATION CODE: DUI
ARREST TIME: 01:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: ROBINSON

DATE/TIME OF TEST: 12/25/2017 01:56:50

Diagnostics Check	OK	01:50
Air Blank	0.000	01:50
Control Test	0.081	01:51
Air Blank	0.000	01:51
Subject Sample # 1	0.261	01:52
Breath Volume	1.320	01:52
Air Blank	0.000	01:52
Air Blank	0.000	01:54
Subject Sample # 2	0.264	01:55
Breath Volume	2.191	01:55
Air Blank	0.000	01:55
Control Test	0.082	01:56
Air Blank	0.000	01:56
Diagnostics Check	OK	01:56

Data Download Date/Time: 12/28/2017 12:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001155

INSTRUMENT REGISTERED TO: EGLIN AFB
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMIH, TYLER, L
BREATH TEST OPERATOR AGENCY: Eglin AFB PD
ARRESTING OFFICER: SMITH, TYLER
ARRESTING OFFICER AGENCY: Eglin AFB PD
LAST AGENCY INSPECTION: 11/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:46

SUBJECT'S LAST NAME: MELTON

DATE/TIME OF TEST: 12/03/2017 02:54:38

Diagnostics Check	OK	02:47
Air Blank	0.000	02:48
Control Test	0.079	02:48
Air Blank	0.000	02:48
Subject Sample # 1	0.179	02:50
Breath Volume	2.066	02:50
Air Blank	0.000	02:50
Air Blank	0.000	02:52
Subject Sample # 2	0.165	02:53
Breath Volume	2.093	02:53
Air Blank	0.000	02:53
Control Test	0.079	02:54
Air Blank	0.000	02:54
Diagnostics Check	OK	02:54

Data Download Date/Time: 12/13/2017 14:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001156

INSTRUMENT REGISTERED TO: ORMOND BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUBIN,RICHARD,D
BREATH TEST OPERATOR AGENCY: Ormond Beach PD
ARRESTING OFFICER: EDWARDS,L
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 814183
DRY GAS EXPIRY DATE: 07/26/2020
VIOLATION CODE: DUI
ARREST TIME: 02:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: SMALL

DATE/TIME OF TEST: 12/03/2017 03:55:13

Diagnostics Check	OK	03:52	
Air Blank	0.000	03:52	
Control Test	0.080	03:53	
Air Blank	0.000	03:53	
Subject Sample	# 1	03:53	Subject Test Refused
Breath Volume	0.000	03:53	
Air Blank	0.000	03:54	
Control Test	0.081	03:54	
Air Blank	0.000	03:55	
Diagnostics Check	OK	03:55	

Data Download Date/Time: 12/26/2017 11:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001156

INSTRUMENT REGISTERED TO: ORMOND BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMILTON,BENITA,L
BREATH TEST OPERATOR AGENCY: Ormond Beach PD
ARRESTING OFFICER: RAIHALA,JON
ARRESTING OFFICER AGENCY: Ormond Beach PD
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 814183
DRY GAS EXPIRY DATE: 07/26/2020
VIOLATION CODE: DUI
ARREST TIME: 10:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:10

SUBJECT'S LAST NAME: BRENEMAN

DATE/TIME OF TEST: 12/07/2017 11:38:15

Diagnostics Check	OK	11:30
Air Blank	0.000	11:31
Control Test	0.081	11:31
Air Blank	0.000	11:32
Subject Sample # 1	0.297	11:33
Breath Volume	3.546	11:33
Air Blank	0.000	11:34
Air Blank	0.000	11:36
Subject Sample # 2	0.279	11:36
Breath Volume	2.527	11:36
Air Blank	0.000	11:37
Control Test	0.079	11:37
Air Blank	0.000	11:38
Diagnostics Check	OK	11:38

Data Download Date/Time: 12/26/2017 11:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001156

INSTRUMENT REGISTERED TO: ORMOND BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEARSON,ROBERT,B
BREATH TEST OPERATOR AGENCY: Ormond Beach PD
ARRESTING OFFICER: RUBIN,RICHARD
ARRESTING OFFICER AGENCY: Ormond Beach PD
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 814183
DRY GAS EXPIRY DATE: 07/26/2020
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: RUEDA JR

DATE/TIME OF TEST: 12/18/2017 03:50:38

Diagnostics Check	OK	03:43
Air Blank	0.000	03:44
Control Test	0.081	03:44
Air Blank	0.000	03:45
Subject Sample # 1	0.207	03:46
Breath Volume	3.843	03:46
Air Blank	0.000	03:46
Air Blank	0.000	03:48
Subject Sample # 2	0.188	03:49
Breath Volume	2.914	03:49
Air Blank	0.000	03:49
Control Test	0.080	03:50
Air Blank	0.000	03:50
Diagnostics Check	OK	03:50

Data Download Date/Time: 12/26/2017 11:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001156

INSTRUMENT REGISTERED TO: ORMOND BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARMINGTON, WILLIAM, G
BREATH TEST OPERATOR AGENCY: Ormond Beach PD
ARRESTING OFFICER: DOUGLAS, THOMAS
ARRESTING OFFICER AGENCY: Ormond Beach PD
LAST AGENCY INSPECTION: 12/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 814183
DRY GAS EXPIRY DATE: 07/26/2020
VIOLATION CODE: DUI
ARREST TIME: 01:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: ELLIS

DATE/TIME OF TEST: 12/29/2017 03:04:27

Diagnostics Check	OK	02:58
Air Blank	0.000	02:58
Control Test	0.081	02:58
Air Blank	0.000	02:59
Subject Sample # 1	0.170	02:59
Breath Volume	2.871	02:59
Air Blank	0.000	03:00
Air Blank	0.000	03:02
Subject Sample # 2	0.176	03:02
Breath Volume	2.816	03:02
Air Blank	0.000	03:03
Control Test	0.081	03:03
Air Blank	0.000	03:04
Diagnostics Check	OK	03:04

Data Download Date/Time: 01/27/2018 18:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001158

INSTRUMENT REGISTERED TO: PONCE INLET PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEAD,C,A
BREATH TEST OPERATOR AGENCY: Ponce Inlet PD
ARRESTING OFFICER: MEAD,C
ARRESTING OFFICER AGENCY: Ponce Inlet PD
LAST AGENCY INSPECTION: 12/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 814183
DRY GAS EXPIRY DATE: 07/26/2020
VIOLATION CODE: DUI
ARREST TIME: 23:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:54

SUBJECT'S LAST NAME: COLBY

DATE/TIME OF TEST: 12/23/2017 00:30:21

Diagnostics Check	OK	00:23
Air Blank	0.000	00:23
Control Test	0.082	00:23
Air Blank	0.000	00:24
Subject Sample # 1	0.229	00:25
Breath Volume	2.144	00:25
Air Blank	0.000	00:26
Air Blank	0.000	00:28
Subject Sample # 2	0.230	00:28
Breath Volume	2.234	00:28
Air Blank	0.000	00:29
Control Test	0.082	00:29
Air Blank	0.000	00:30
Diagnostics Check	OK	00:30

Data Download Date/Time: 01/20/2018 01:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001159

INSTRUMENT REGISTERED TO: ORANGE CITY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRUBAKER,KELLY,J
BREATH TEST OPERATOR AGENCY: Orange City PD
ARRESTING OFFICER: MAGER,CHRISTOPHER
ARRESTING OFFICER AGENCY: Orange City PD
LAST AGENCY INSPECTION: 12/16/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1731608A2
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 15:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:30

SUBJECT'S LAST NAME: RIVERA ORTIZ

DATE/TIME OF TEST: 12/25/2017 16:03:17

Diagnostics Check	OK	15:56
Air Blank	0.000	15:56
Control Test	0.080	15:57
Air Blank	0.000	15:57
Subject Sample # 1	0.162	15:58
Breath Volume	1.324	15:58
Air Blank	0.000	15:59
Air Blank	0.000	16:00
Subject Sample # 2	0.154	16:01
Breath Volume	2.035	16:01
Air Blank	0.000	16:02
Control Test	0.080	16:02
Air Blank	0.000	16:03
Diagnostics Check	OK	16:03

Data Download Date/Time: 01/14/2018 14:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001190

INSTRUMENT REGISTERED TO: BOYNTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASTRO,DENNIS,
BREATH TEST OPERATOR AGENCY: Boynton Beach PD
ARRESTING OFFICER: STRONG,KENDRA
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 11/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 390314
DRY GAS EXPIRY DATE: 04/19/2019
VIOLATION CODE: DUI
ARREST TIME: 19:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:53

SUBJECT'S LAST NAME: DURHAM

DATE/TIME OF TEST: 12/09/2017 20:52:43

Diagnostics Check	OK	20:46
Air Blank	0.000	20:46
Control Test	0.079	20:47
Air Blank	0.000	20:47
Subject Sample # 1	0.106	20:48
Breath Volume	2.582	20:48
Air Blank	0.000	20:48
Air Blank	0.000	20:50
Subject Sample # 2	0.106	20:51
Breath Volume	2.976	20:51
Air Blank	0.000	20:51
Control Test	0.080	20:52
Air Blank	0.000	20:52
Diagnostics Check	OK	20:52

Data Download Date/Time: 12/13/2017 21:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001190

INSTRUMENT REGISTERED TO: BOYNTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASTRO,DENNIS,
BREATH TEST OPERATOR AGENCY: Boynton Beach PD
ARRESTING OFFICER: CASTRO,DENNIS
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 12/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 390314
DRY GAS EXPIRY DATE: 04/19/2019
VIOLATION CODE: DUI
ARREST TIME: 22:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:32

SUBJECT'S LAST NAME: CHERISCART

DATE/TIME OF TEST: 12/15/2017 22:59:40

Diagnostics Check	OK	22:53
Air Blank	0.000	22:53
Control Test	0.081	22:54
Air Blank	0.000	22:54
Subject Sample # 1	0.000	22:55
Breath Volume	1.988	22:55
Air Blank	0.000	22:55
Air Blank	0.000	22:57
Subject Sample # 2	0.000	22:58
Breath Volume	2.425	22:58
Air Blank	0.000	22:58
Control Test	0.081	22:59
Air Blank	0.000	22:59
Diagnostics Check	OK	22:59

Data Download Date/Time: 01/27/2018 17:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001190

INSTRUMENT REGISTERED TO: BOYNTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MUNRO,CHRISTOPHER,A
BREATH TEST OPERATOR AGENCY: Boynton Beach PD
ARRESTING OFFICER: MUNRO,CHRISTOPHER
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 12/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 390314
DRY GAS EXPIRY DATE: 04/19/2019
VIOLATION CODE: DUI
ARREST TIME: 13:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:49

SUBJECT'S LAST NAME: EATON

DATE/TIME OF TEST: 12/19/2017 15:23:25

Diagnostics Check	OK	15:17
Air Blank	0.000	15:17
Control Test	0.080	15:17
Air Blank	0.000	15:18
Subject Sample # 1	0.000	15:19
Breath Volume	2.152	15:19
Air Blank	0.000	15:19
Air Blank	0.000	15:21
Subject Sample # 2	0.000	15:22
Breath Volume	1.546	15:22
Air Blank	0.000	15:22
Control Test	0.080	15:22
Air Blank	0.000	15:23
Diagnostics Check	OK	15:23

Data Download Date/Time: 01/27/2018 17:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001201

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOPEZ,JOSE,D
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: LOPEZ,JOSE
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 11/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A2
DRY GAS EXPIRY DATE: 03/06/2019
VIOLATION CODE: DUI
ARREST TIME: 00:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:04

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 12/03/2017 01:36:43

Diagnostics Check	OK	01:30
Air Blank	0.000	01:30
Control Test	0.079	01:31
Air Blank	0.000	01:31
Subject Sample # 1	0.203	01:32
Breath Volume	3.695	01:32
Air Blank	0.000	01:32
Air Blank	0.000	01:34
Subject Sample # 2	0.201	01:35
Breath Volume	3.710	01:35
Air Blank	0.000	01:35
Control Test	0.080	01:36
Air Blank	0.000	01:36
Diagnostics Check	OK	01:36

Data Download Date/Time: 12/06/2017 12:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001201

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOPEZ,JACOB,
BREATH TEST OPERATOR AGENCY: Seminole Dept of LE
ARRESTING OFFICER: HERRON,LAURA
ARRESTING OFFICER AGENCY: Seminole Dept of LE
LAST AGENCY INSPECTION: 11/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A2
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 07:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:07

SUBJECT'S LAST NAME: GARCIAGABRIEL

DATE/TIME OF TEST: 12/03/2017 08:43:35

Diagnostics Check	OK	08:37
Air Blank	0.000	08:37
Control Test	0.079	08:38
Air Blank	0.000	08:38
Subject Sample # 1	0.171	08:39
Breath Volume	3.707	08:39
Air Blank	0.000	08:39
Air Blank	0.000	08:41
Subject Sample # 2	0.167	08:42
Breath Volume	2.226	08:42
Air Blank	0.000	08:42
Control Test	0.079	08:43
Air Blank	0.000	08:43
Diagnostics Check	OK	08:43

Data Download Date/Time: 12/06/2017 12:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAND,CURTIS,L
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: BLAND,CURTIS
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 12/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 366849
DRY GAS EXPIRY DATE: 03/30/2019
VIOLATION CODE: DUI
ARREST TIME: 01:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: LOPEZ COLON

DATE/TIME OF TEST: 12/10/2017 02:18:24

Diagnostics Check		OK	02:11
Air Blank		0.000	02:11
Control Test		0.080	02:12
Air Blank		0.000	02:12
Subject Sample	# 1	0.249	02:13
Breath Volume		1.296	02:13
Air Blank		0.000	02:14
Air Blank		0.000	02:16
Subject Sample	# 2	0.257	02:16
Breath Volume		2.953	02:16
Air Blank		0.000	02:17
Control Test		0.079	02:17
Air Blank		0.000	02:18
Diagnostics Check		OK	02:18

Data Download Date/Time: 01/30/2018 12:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAND,CURTIS,L
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: NELSON,GARRETT
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 12/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 366849
DRY GAS EXPIRY DATE: 03/30/2019
VIOLATION CODE: DUI
ARREST TIME: 04:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:15

SUBJECT'S LAST NAME: CRUTCHFIELD

DATE/TIME OF TEST: 12/11/2017 04:43:12

Diagnostics Check	OK	04:36
Air Blank	0.000	04:36
Control Test	0.081	04:37
Air Blank	0.000	04:37
Subject Sample # 1	0.105	04:38
Breath Volume	1.394	04:38
Air Blank	0.000	04:39
Air Blank	0.000	04:40
Subject Sample # 2	0.102	04:41
Breath Volume	3.507	04:41
Air Blank	0.000	04:42
Control Test	0.080	04:42
Air Blank	0.000	04:43
Diagnostics Check	OK	04:43

Data Download Date/Time: 01/30/2018 12:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROGERS,JOHN,J
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: BOHRER,MARK
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 12/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 366849
DRY GAS EXPIRY DATE: 03/30/2019
VIOLATION CODE: DUI
ARREST TIME: 02:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: DELGADO

DATE/TIME OF TEST: 12/17/2017 03:28:02

Diagnostics Check	OK	03:20
Air Blank	0.000	03:21
Control Test	0.081	03:21
Air Blank	0.000	03:22
Subject Sample # 1	0.018	03:22
Breath Volume	1.589	03:22
Air Blank	0.000	03:23
Air Blank	0.000	03:25
Subject Sample # 2	0.017	03:26
Breath Volume	1.867	03:26
Air Blank	0.000	03:27
Control Test	0.081	03:27
Air Blank	0.000	03:27
Diagnostics Check	OK	03:28

Data Download Date/Time: 01/30/2018 12:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLORES,BARRY,G
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: FLORES,BARRY
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 12/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 366849
DRY GAS EXPIRY DATE: 03/30/2019
VIOLATION CODE: DUI
ARREST TIME: 20:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:45

SUBJECT'S LAST NAME: JIMENEZ

DATE/TIME OF TEST: 12/17/2017 21:13:21

Diagnostics Check	OK	21:05
Air Blank	0.000	21:06
Control Test	0.077	21:06
Air Blank	0.000	21:06
Subject Sample # 1	0.180	21:08
Breath Volume	1.554	21:08
Air Blank	0.000	21:08
Air Blank	0.000	21:10
Subject Sample # 2	0.178	21:11
Breath Volume	1.355	21:11
Air Blank	0.000	21:12
Control Test	0.077	21:12
Air Blank	0.000	21:13
Diagnostics Check	OK	21:13

Data Download Date/Time: 01/30/2018 12:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAND,CURTIS,L
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: NEESE,CORBIN
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 12/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 366849
DRY GAS EXPIRY DATE: 03/30/2019
VIOLATION CODE: DUI
ARREST TIME: 00:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:18

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 12/19/2017 00:45:36

Diagnostics Check		OK	00:39
Air Blank		0.000	00:39
Control Test		0.080	00:40
Air Blank		0.000	00:40
Subject Sample	# 1	0.174	00:41
Breath Volume		1.531	00:41
Air Blank		0.000	00:41
Air Blank		0.000	00:43
Subject Sample	# 2	0.180	00:44
Breath Volume		1.406	00:44
Air Blank		0.000	00:44
Control Test		0.080	00:45
Air Blank		0.000	00:45
Diagnostics Check		OK	00:45

Data Download Date/Time: 01/30/2018 12:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAND,CURTIS,L
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: NELSON,GARRETT
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 12/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 366849
DRY GAS EXPIRY DATE: 03/30/2019
VIOLATION CODE: DUI
ARREST TIME: 22:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: WIMBERLY

DATE/TIME OF TEST: 12/21/2017 23:28:15

Diagnostics Check	OK	23:20
Air Blank	0.000	23:20
Control Test	0.080	23:21
Air Blank	0.000	23:21
Subject Sample # 1	0.195	23:23
Breath Volume	1.609	23:23
Air Blank	0.000	23:24
Air Blank	0.000	23:25
Subject Sample # 2	0.197	23:26
Breath Volume	1.960	23:26
Air Blank	0.000	23:27
Control Test	0.080	23:27
Air Blank	0.000	23:28
Diagnostics Check	OK	23:28

Data Download Date/Time: 01/30/2018 12:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAND,CURTIS,L
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: BLAND,CURTIS
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 12/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 366849
DRY GAS EXPIRY DATE: 03/30/2019
VIOLATION CODE: DUI
ARREST TIME: 01:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: MERRISS

DATE/TIME OF TEST: 12/23/2017 01:57:24

Diagnostics Check	OK	01:50
Air Blank	0.000	01:51
Control Test	0.080	01:51
Air Blank	0.000	01:52
Subject Sample # 1	0.361	01:52
Breath Volume	3.316	01:52
Air Blank	0.000	01:53
Air Blank	0.000	01:55
Subject Sample # 2	0.356	01:55
Breath Volume	4.496	01:55
Air Blank	0.000	01:56
Control Test	0.080	01:56
Air Blank	0.000	01:57
Diagnostics Check	OK	01:57

Data Download Date/Time: 01/30/2018 12:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAND,CURTIS,L
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: BROCK,AUSTIN
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 12/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 366849
DRY GAS EXPIRY DATE: 03/30/2019
VIOLATION CODE: DUI
ARREST TIME: 05:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:13

SUBJECT'S LAST NAME: COOK

DATE/TIME OF TEST: 12/23/2017 05:40:21

Diagnostics Check	OK	05:33
Air Blank	0.000	05:34
Control Test	0.081	05:34
Air Blank	0.000	05:34
Subject Sample # 1	0.198	05:35
Breath Volume	1.718	05:35
Air Blank	0.000	05:36
Air Blank	0.000	05:38
Subject Sample # 2	0.195	05:38
Breath Volume	3.250	05:38
Air Blank	0.000	05:39
Control Test	0.080	05:39
Air Blank	0.000	05:40
Diagnostics Check	OK	05:40

Data Download Date/Time: 01/30/2018 12:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001207

INSTRUMENT REGISTERED TO: HENDRY COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCLAIN,KYLE,J
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: MCLAIN,KYLE
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 12/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 83551
DRY GAS EXPIRY DATE: 10/02/2018
VIOLATION CODE: DUI
ARREST TIME: 02:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: MANKO

DATE/TIME OF TEST: 12/30/2017 03:22:34

Diagnostics Check	OK	03:15
Air Blank	0.000	03:16
Control Test	0.081	03:16
Air Blank	0.000	03:17
Subject Sample # 1	0.120	03:17
Breath Volume	2.554	03:17
Air Blank	0.000	03:18
Air Blank	0.000	03:20
Subject Sample # 2	0.115	03:21
Breath Volume	2.593	03:21
Air Blank	0.000	03:21
Control Test	0.081	03:21
Air Blank	0.000	03:22
Diagnostics Check	OK	03:22

Data Download Date/Time: 01/23/2018 22:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001207

INSTRUMENT REGISTERED TO: HENDRY COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NORWOOD,CHRISTOPHER,J
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: WOOSLEY,LAWRENCE
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 12/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 83551
DRY GAS EXPIRY DATE: 10/02/2018
VIOLATION CODE: DUI
ARREST TIME: 16:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:00

SUBJECT'S LAST NAME: PALACIOS FLORES

DATE/TIME OF TEST: 12/30/2017 17:29:28

Diagnostics Check	OK	17:20
Air Blank	0.000	17:21
Control Test	0.082	17:21
Air Blank	0.000	17:21
Subject Sample # 1	0.243	17:24
Breath Volume	1.843	17:24
Air Blank	0.000	17:24
Air Blank	0.000	17:26
Subject Sample # 2	0.238	17:27
Breath Volume	2.250	17:27
Air Blank	0.000	17:28
Control Test	0.082	17:28
Air Blank	0.000	17:29
Diagnostics Check	OK	17:29

Data Download Date/Time: 01/23/2018 22:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001230

INSTRUMENT REGISTERED TO: TYNDALL AFB
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBER,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Tyndall AFB PD
ARRESTING OFFICER: WEBER,JEFFREY
ARRESTING OFFICER AGENCY: Tyndall AFB PD
LAST AGENCY INSPECTION: 12/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:09

SUBJECT'S LAST NAME: BUSH

DATE/TIME OF TEST: 12/09/2017 01:36:18

Diagnostics Check	OK	01:30
Air Blank	0.000	01:30
Control Test	0.080	01:30
Air Blank	0.000	01:31
Subject Sample # 1	0.000	01:32
Breath Volume	1.976	01:32
Air Blank	0.000	01:32
Air Blank	0.000	01:34
Subject Sample # 2	0.000	01:34
Breath Volume	2.187	01:34
Air Blank	0.000	01:35
Control Test	0.080	01:35
Air Blank	0.000	01:36
Diagnostics Check	OK	01:36

Data Download Date/Time: 01/03/2018 12:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001230

INSTRUMENT REGISTERED TO: TYNDALL AFB
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PLACE,NICHOLAS,T
BREATH TEST OPERATOR AGENCY: Tyndall AFB PD
ARRESTING OFFICER: WEBER,JEFFREY
ARRESTING OFFICER AGENCY: Tyndall AFB PD
LAST AGENCY INSPECTION: 12/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:54

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 12/10/2017 03:23:01

Diagnostics Check	OK	03:14
Air Blank	0.000	03:14
Control Test	0.082	03:15
Air Blank	0.000	03:15
Subject Sample # 1	0.058	03:18
Breath Volume	1.140	03:18
Air Blank	0.000	03:19
Air Blank	0.000	03:20
Subject Sample # 2	0.060	03:21
Breath Volume	1.554	03:21
Air Blank	0.000	03:22
Control Test	0.081	03:22
Air Blank	0.000	03:22
Diagnostics Check	OK	03:23

Data Download Date/Time: 01/03/2018 12:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001230

INSTRUMENT REGISTERED TO: TYNDALL AFB
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBER,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Tyndall AFB PD
ARRESTING OFFICER: STRICKLAND,MIKAELA
ARRESTING OFFICER AGENCY: Parker PD
LAST AGENCY INSPECTION: 12/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:51

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 12/11/2017 02:26:48

Diagnostics Check	OK	02:20
Air Blank	0.000	02:21
Control Test	0.081	02:21
Air Blank	0.000	02:21
Subject Sample # 1	0.108	02:22
Breath Volume	3.574	02:22
Air Blank	0.000	02:22
Air Blank	0.000	02:24
Subject Sample # 2	0.111	02:25
Breath Volume	2.335	02:25
Air Blank	0.000	02:25
Control Test	0.080	02:26
Air Blank	0.000	02:26
Diagnostics Check	OK	02:26

Data Download Date/Time: 01/03/2018 12:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001230

INSTRUMENT REGISTERED TO: TYNDALL AFB
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARZA,IVAN,A
BREATH TEST OPERATOR AGENCY: Tyndall AFB PD
ARRESTING OFFICER: GARZA,IVAN
ARRESTING OFFICER AGENCY: Tyndall AFB PD
LAST AGENCY INSPECTION: 12/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: ADMIN
ARREST TIME: 15:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:23

SUBJECT'S LAST NAME: BUCKMAN

DATE/TIME OF TEST: 12/18/2017 15:56:12

Diagnostics Check	OK	15:49
Air Blank	0.000	15:49
Control Test	0.079	15:50
Air Blank	0.000	15:50
Subject Sample # 1	0.289	15:51
Breath Volume	1.214	15:51
Air Blank	0.000	15:51
Air Blank	0.000	15:53
Subject Sample # 2	0.275	15:54
Breath Volume	1.226	15:54
Air Blank	0.000	15:55
Control Test	0.078	15:55
Air Blank	0.000	15:56
Diagnostics Check	OK	15:56

Data Download Date/Time: 01/03/2018 12:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001230

INSTRUMENT REGISTERED TO: TYNDALL AFB
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PLACE,NICHOLAS,T
BREATH TEST OPERATOR AGENCY: Tyndall AFB PD
ARRESTING OFFICER: TAYLOR,CHARLES
ARRESTING OFFICER AGENCY: Tyndall AFB PD
LAST AGENCY INSPECTION: 12/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 19:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:11

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 12/20/2017 19:37:35

Diagnostics Check	OK	19:31
Air Blank	0.000	19:32
Control Test	0.080	19:32
Air Blank	0.000	19:32
Subject Sample # 1	0.000	19:33
Breath Volume	1.511	19:33
Air Blank	0.000	19:33
Air Blank	0.000	19:35
Subject Sample # 2	0.000	19:36
Breath Volume	1.519	19:36
Air Blank	0.000	19:36
Control Test	0.080	19:36
Air Blank	0.000	19:37
Diagnostics Check	OK	19:37

Data Download Date/Time: 01/03/2018 12:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001230

INSTRUMENT REGISTERED TO: TYNDALL AFB
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PLACE,NICHOLAS,T
BREATH TEST OPERATOR AGENCY: Tyndall AFB PD
ARRESTING OFFICER: WEBER,JEFFERY
ARRESTING OFFICER AGENCY: Tyndall AFB PD
LAST AGENCY INSPECTION: 12/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:39

SUBJECT'S LAST NAME: RAMIREZ

DATE/TIME OF TEST: 12/27/2017 21:05:53

Diagnostics Check	OK	20:59
Air Blank	0.000	21:00
Control Test	0.081	21:00
Air Blank	0.000	21:00
Subject Sample # 1	0.064	21:01
Breath Volume	2.773	21:01
Air Blank	0.000	21:02
Air Blank	0.000	21:03
Subject Sample # 2	0.065	21:04
Breath Volume	1.667	21:04
Air Blank	0.000	21:04
Control Test	0.081	21:05
Air Blank	0.000	21:05
Diagnostics Check	OK	21:05

Data Download Date/Time: 01/03/2018 12:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001230

INSTRUMENT REGISTERED TO: TYNDALL AFB
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORGAN,AQUILA,A
BREATH TEST OPERATOR AGENCY: Tyndall AFB PD
ARRESTING OFFICER: TAYLOR,CHARLES
ARRESTING OFFICER AGENCY: Tyndall AFB PD
LAST AGENCY INSPECTION: 12/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:11

SUBJECT'S LAST NAME: SHOCKLEY

DATE/TIME OF TEST: 12/30/2017 04:38:08

Diagnostics Check	OK	04:31
Air Blank	0.000	04:32
Control Test	0.081	04:32
Air Blank	0.000	04:32
Subject Sample # 1	0.000	04:33
Breath Volume	1.417	04:33
Air Blank	0.000	04:34
Air Blank	0.000	04:36
Subject Sample # 2	0.000	04:36
Breath Volume	1.675	04:36
Air Blank	0.000	04:37
Control Test	0.081	04:37
Air Blank	0.000	04:38
Diagnostics Check	OK	04:38

Data Download Date/Time: 01/03/2018 12:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JEUDY,RINALDO,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: WOFFORD,CLINT
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/23/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: RADFORD

DATE/TIME OF TEST: 12/24/2017 00:35:12

Diagnostics Check		OK	00:25
Air Blank		0.000	00:25
Control Test		0.079	00:25
Air Blank		0.000	00:26
Subject Sample	# 1	0.284	00:27
Breath Volume		4.167	00:27
Air Blank		0.000	00:27
Air Blank		0.000	00:29
Subject Sample	# 2	0.261	00:30
Breath Volume		2.812	00:30
Air Blank		0.000	00:31
Air Blank		0.000	00:32
Subject Sample	# 3	0.273	00:33
Breath Volume		3.593	00:33
Air Blank		0.000	00:34
Control Test		0.078	00:34
Air Blank		0.000	00:35
Diagnostics Check		OK	00:35

Data Download Date/Time: 01/20/2018 19:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUCAS,NATHANIEL,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: LUCAS,NATHANIEL
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/23/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: BETANCOURT

DATE/TIME OF TEST: 12/24/2017 03:43:30

Diagnostics Check		OK	03:32
Air Blank		0.000	03:32
Control Test		0.079	03:33
Air Blank		0.000	03:33
Subject Sample	# 1	0.287	03:35
Breath Volume		2.019	03:35
Air Blank		0.000	03:36
Air Blank		0.000	03:38
Subject Sample	# 2	0.311	03:38
Breath Volume		1.285	03:38
Air Blank		0.000	03:39
Air Blank		0.000	03:41
Subject Sample	# 3	0.316	03:41
Breath Volume		1.691	03:41
Air Blank		0.000	03:42
Control Test		0.077	03:42
Air Blank		0.000	03:43
Diagnostics Check		OK	03:43

Data Download Date/Time: 01/20/2018 19:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUCAS,NATHANIEL,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: WOFFORD,CLINT
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/23/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:02

SUBJECT'S LAST NAME: VARNADOE

DATE/TIME OF TEST: 12/24/2017 20:42:03

Diagnostics Check	OK	20:35
Air Blank	0.000	20:36
Control Test	0.080	20:36
Air Blank	0.000	20:37
Subject Sample # 1	0.000	20:37
Breath Volume	2.539	20:37
Air Blank	0.000	20:38
Air Blank	0.000	20:40
Subject Sample # 2	0.000	20:40
Breath Volume	1.941	20:40
Air Blank	0.000	20:41
Control Test	0.079	20:41
Air Blank	0.000	20:41
Diagnostics Check	OK	20:42

Data Download Date/Time: 01/20/2018 19:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUCAS,NATHANIEL,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: LUCAS,NATHANIEL
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/23/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:30

SUBJECT'S LAST NAME: RODRIGUEZ EVORA

DATE/TIME OF TEST: 12/24/2017 21:02:00

Diagnostics Check	OK	20:55
Air Blank	0.000	20:55
Control Test	0.080	20:56
Air Blank	0.000	20:56
Subject Sample # 1	0.221	20:57
Breath Volume	2.007	20:57
Air Blank	0.000	20:58
Air Blank	0.000	21:00
Subject Sample # 2	0.227	21:00
Breath Volume	1.867	21:00
Air Blank	0.000	21:01
Control Test	0.079	21:01
Air Blank	0.000	21:01
Diagnostics Check	OK	21:02

Data Download Date/Time: 01/20/2018 19:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PHILLIPS,JORDAN,W
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: WOFFORD,CLINT
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/23/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: GOMSI

DATE/TIME OF TEST: 12/25/2017 03:42:54

Diagnostics Check	OK	03:35
Air Blank	0.000	03:36
Control Test	0.079	03:36
Air Blank	0.000	03:36
Subject Sample # 1	0.070	03:37
Breath Volume	1.570	03:37
Air Blank	0.000	03:38
Air Blank	0.000	03:40
Subject Sample # 2	0.074	03:41
Breath Volume	2.328	03:41
Air Blank	0.000	03:42
Control Test	0.080	03:42
Air Blank	0.000	03:42
Diagnostics Check	OK	03:42

Data Download Date/Time: 01/20/2018 19:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JEUDY,RINALDO,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: CATELLO,MARIA
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/23/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:42

SUBJECT'S LAST NAME: SPRUILL

DATE/TIME OF TEST: 12/25/2017 05:19:51

Diagnostics Check	OK	05:16	
Air Blank	0.000	05:17	
Control Test	0.080	05:17	
Air Blank	0.000	05:18	
Subject Sample	# 1	05:18	Subject Test Refused
Breath Volume	0.000	05:18	
Air Blank	0.000	05:18	
Control Test	0.080	05:19	
Air Blank	0.000	05:19	
Diagnostics Check	OK	05:19	

Data Download Date/Time: 01/20/2018 19:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COSTALES,ALEXANDRO,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: COSTALES,ALEXANDRO
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/23/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 07:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:32

SUBJECT'S LAST NAME: CRUZ PERALTA

DATE/TIME OF TEST: 12/25/2017 08:59:33

Diagnostics Check	OK	08:53
Air Blank	0.000	08:53
Control Test	0.079	08:53
Air Blank	0.000	08:54
Subject Sample # 1	0.126	08:54
Breath Volume	2.789	08:54
Air Blank	0.000	08:55
Air Blank	0.000	08:57
Subject Sample # 2	0.125	08:58
Breath Volume	1.910	08:58
Air Blank	0.000	08:58
Control Test	0.080	08:58
Air Blank	0.000	08:59
Diagnostics Check	OK	08:59

Data Download Date/Time: 01/20/2018 19:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAWES,ERIC,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: LAWES,ERIC
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/23/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 15:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:03

SUBJECT'S LAST NAME: LEWIS

DATE/TIME OF TEST: 12/25/2017 16:40:08

Diagnostics Check	OK	16:33
Air Blank	0.000	16:34
Control Test	0.080	16:34
Air Blank	0.000	16:34
Subject Sample # 1	0.151	16:35
Breath Volume	3.445	16:35
Air Blank	0.000	16:36
Air Blank	0.000	16:38
Subject Sample # 2	0.145	16:38
Breath Volume	3.070	16:38
Air Blank	0.000	16:39
Control Test	0.080	16:39
Air Blank	0.000	16:40
Diagnostics Check	OK	16:40

Data Download Date/Time: 01/20/2018 19:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCINTOSH,HALLIE,T
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: WOFFORD,CLINTON
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/23/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:30

SUBJECT'S LAST NAME: PRUITT

DATE/TIME OF TEST: 12/26/2017 21:07:05

Diagnostics Check	OK	20:58	
Air Blank	0.000	20:58	
Control Test	0.080	20:59	
Air Blank	0.000	20:59	
Subject Sample	# 1 0.363	21:00	
Breath Volume	2.242	21:00	
Air Blank	0.000	21:00	
Air Blank	0.000	21:02	
Subject Sample	# 2 0.272	21:05	Volume Not Met
Breath Volume	0.878	21:05	
Air Blank		21:06	Purge Fail
Air Blank	0.000	21:07	

Data Download Date/Time: 01/20/2018 19:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCINTOSH,HALLIE,T
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: WOFFORD,CLINTON
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/23/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:30

SUBJECT'S LAST NAME: PRUITT

DATE/TIME OF TEST: 12/26/2017 21:15:44

Diagnostics Check	OK	21:11	
Air Blank	0.000	21:11	
Control Test	0.075	21:12	
Air Blank	0.000	21:12	
Subject Sample # 1	0.323	21:14	
Breath Volume	1.714	21:14	
Air Blank		21:15	Purge Fail
Air Blank	0.000	21:15	

Data Download Date/Time: 01/20/2018 19:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCINTOSH,HALLIE,T
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: WOFFORD,CLINT
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/23/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: SCALPI

DATE/TIME OF TEST: 12/26/2017 23:52:35

Diagnostics Check	OK	23:45
Air Blank	0.000	23:46
Control Test	0.079	23:46
Air Blank	0.000	23:47
Subject Sample # 1	0.136	23:47
Breath Volume	2.132	23:47
Air Blank	0.000	23:48
Air Blank	0.000	23:50
Subject Sample # 2	0.126	23:51
Breath Volume	1.269	23:51
Air Blank	0.000	23:51
Control Test	0.079	23:51
Air Blank	0.000	23:52
Diagnostics Check	OK	23:52

Data Download Date/Time: 01/20/2018 19:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUCAS,NATHANIEL,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: LUCAS,NATHANIEL
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/23/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: BENEFIELD

DATE/TIME OF TEST: 12/27/2017 00:23:14

Diagnostics Check	OK	00:16
Air Blank	0.000	00:16
Control Test	0.079	00:16
Air Blank	0.000	00:17
Subject Sample # 1	0.205	00:18
Breath Volume	1.628	00:18
Air Blank	0.000	00:19
Air Blank	0.000	00:21
Subject Sample # 2	0.201	00:21
Breath Volume	1.496	00:21
Air Blank	0.000	00:22
Control Test	0.078	00:22
Air Blank	0.000	00:23
Diagnostics Check	OK	00:23

Data Download Date/Time: 01/20/2018 19:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JEUDY,RINALDO,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: WOFFORD,CLINT
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/23/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:21

SUBJECT'S LAST NAME: SPAULDING

DATE/TIME OF TEST: 12/28/2017 21:56:33

Diagnostics Check	OK	21:49
Air Blank	0.000	21:50
Control Test	0.080	21:50
Air Blank	0.000	21:50
Subject Sample # 1	0.075	21:51
Breath Volume	2.417	21:51
Air Blank	0.000	21:52
Air Blank	0.000	21:54
Subject Sample # 2	0.075	21:55
Breath Volume	2.335	21:55
Air Blank	0.000	21:55
Control Test	0.080	21:55
Air Blank	0.000	21:56
Diagnostics Check	OK	21:56

Data Download Date/Time: 01/20/2018 19:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GETSEE,JOY,S
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: WOFFORD,CLINT
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/23/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:12

SUBJECT'S LAST NAME: BRAIL

DATE/TIME OF TEST: 12/29/2017 02:55:42

Diagnostics Check	OK	02:49
Air Blank	0.000	02:49
Control Test	0.079	02:50
Air Blank	0.000	02:50
Subject Sample # 1	0.121	02:51
Breath Volume	2.109	02:51
Air Blank	0.000	02:51
Air Blank	0.000	02:53
Subject Sample # 2	0.124	02:54
Breath Volume	2.132	02:54
Air Blank	0.000	02:54
Control Test	0.080	02:55
Air Blank	0.000	02:55
Diagnostics Check	OK	02:55

Data Download Date/Time: 01/20/2018 19:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAWES,ERIC,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: LAWES,ERIC
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/23/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:32

SUBJECT'S LAST NAME: WHEELER

DATE/TIME OF TEST: 12/29/2017 22:16:04

Diagnostics Check	OK	22:04	
Air Blank	0.000	22:04	
Control Test	0.080	22:05	
Air Blank	0.000	22:05	
Subject Sample	# 1	22:08	No Sample Provided
Breath Volume	0.000	22:08	
Air Blank	0.000	22:09	
Air Blank	0.000	22:11	
Subject Sample	# 2 0.210	22:14	Volume Not Met
Breath Volume	0.273	22:14	
Air Blank	0.000	22:15	
Control Test	0.080	22:15	
Air Blank	0.000	22:15	
Diagnostics Check	OK	22:16	

Data Download Date/Time: 01/20/2018 19:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAWES,ERIC,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: LAWES,ERIC
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/23/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:32

SUBJECT'S LAST NAME: WHEELER

DATE/TIME OF TEST: 12/29/2017 22:40:04

Diagnostics Check		OK	22:23	
Air Blank		0.000	22:24	
Control Test		0.078	22:24	
Air Blank		0.000	22:24	
Subject Sample	# 1	0.238	22:27	No .020 agreement
Breath Volume		1.226	22:27	
Air Blank		0.000	22:28	
Air Blank		0.000	22:29	
Subject Sample	# 2	0.214	22:32	No .020 agreement
Breath Volume		1.812	22:32	
Air Blank		0.000	22:33	
Air Blank		0.000	22:35	
Subject Sample	# 3	0.190	22:38	Volume Not Met
Breath Volume		0.230	22:38	
Air Blank		0.000	22:39	
Control Test		0.075	22:39	
Air Blank		0.000	22:39	
Diagnostics Check		OK	22:40	

Data Download Date/Time: 01/20/2018 19:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SINCLAIR,CAMERON,W
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: SINCLAIR,CAMERON
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/23/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:11

SUBJECT'S LAST NAME: HENAO

DATE/TIME OF TEST: 12/30/2017 22:44:53

Diagnostics Check	OK	22:37
Air Blank	0.000	22:38
Control Test	0.079	22:38
Air Blank	0.000	22:38
Subject Sample # 1	0.210	22:40
Breath Volume	1.160	22:40
Air Blank	0.000	22:41
Air Blank	0.000	22:42
Subject Sample # 2	0.200	22:43
Breath Volume	1.875	22:43
Air Blank	0.000	22:43
Control Test	0.079	22:44
Air Blank	0.000	22:44
Diagnostics Check	OK	22:44

Data Download Date/Time: 01/20/2018 19:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARREN,BILLY,J
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: WARREN,BILLY
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/23/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:37

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 12/31/2017 01:16:17

Diagnostics Check		OK	01:04	
Air Blank		0.000	01:05	
Control Test		0.080	01:05	
Air Blank		0.000	01:05	
Subject Sample	# 1	0.132	01:09	Volume Not Met
Breath Volume		0.054	01:09	
Air Blank		0.000	01:09	
Air Blank		0.000	01:11	
Subject Sample	# 2	0.139	01:14	Volume Not Met
Breath Volume		0.042	01:14	
Air Blank		0.000	01:15	
Control Test		0.078	01:15	
Air Blank		0.000	01:16	
Diagnostics Check		OK	01:16	

Data Download Date/Time: 01/20/2018 19:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAWES,ERIC,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: LAWES,ERIC
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/23/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:09

SUBJECT'S LAST NAME: LESTER

DATE/TIME OF TEST: 12/31/2017 22:38:57

Diagnostics Check	OK	22:35	
Air Blank	0.000	22:36	
Control Test	0.080	22:36	
Air Blank	0.000	22:37	
Subject Sample	# 1	22:37	Interferent Detect
Breath Volume	4.246	22:37	
Air Blank	0.000	22:38	
Control Test	0.080	22:38	
Air Blank	0.000	22:38	
Diagnostics Check	OK	22:38	

Data Download Date/Time: 01/20/2018 19:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAWES,ERIC,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: LAWES,ERIC
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/23/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:09

SUBJECT'S LAST NAME: LESTER

DATE/TIME OF TEST: 12/31/2017 22:47:30

Diagnostics Check	OK	22:41
Air Blank	0.000	22:41
Control Test	0.080	22:42
Air Blank	0.000	22:42
Subject Sample # 1	0.017	22:43
Breath Volume	3.179	22:43
Air Blank	0.000	22:43
Air Blank	0.000	22:45
Subject Sample # 2	0.018	22:46
Breath Volume	5.097	22:46
Air Blank	0.000	22:46
Control Test	0.079	22:46
Air Blank	0.000	22:47
Diagnostics Check	OK	22:47

Data Download Date/Time: 01/20/2018 19:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001238

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PITCHER,DARRYL,P
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: CHESSHER,SANDI
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 12/09/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: KARIM

DATE/TIME OF TEST: 12/28/2017 03:27:03

Diagnostics Check	OK	03:19
Air Blank	0.000	03:20
Control Test	0.079	03:20
Air Blank	0.000	03:21
Subject Sample # 1	0.022	03:22
Breath Volume	1.718	03:22
Air Blank	0.000	03:23
Air Blank	0.000	03:24
Subject Sample # 2	0.022	03:25
Breath Volume	1.101	03:25
Air Blank	0.000	03:26
Control Test	0.079	03:26
Air Blank	0.000	03:26
Diagnostics Check	OK	03:27

Data Download Date/Time: 01/04/2018 17:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001238

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PITCHER,DARRYL,P
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: CHESSHER,SANDRA
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 12/09/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: WELLS

DATE/TIME OF TEST: 12/30/2017 01:20:36

Diagnostics Check	OK	01:12	
Air Blank	0.000	01:12	
Control Test	0.079	01:12	
Air Blank	0.000	01:13	
Subject Sample	# 1	01:16	No Sample Provided
Breath Volume	0.000	01:16	
Air Blank	0.000	01:17	
Air Blank	0.000	01:19	
Subject Sample	# 2	01:19	Subject Test Refused
Breath Volume	0.000	01:19	
Air Blank	0.000	01:19	
Control Test	0.079	01:19	
Air Blank	0.000	01:20	
Diagnostics Check	OK	01:20	

Data Download Date/Time: 01/04/2018 17:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001242

INSTRUMENT REGISTERED TO: DAYTONA BEACH SHORES
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORIARTY, MATTHEW, J
BREATH TEST OPERATOR AGENCY: Daytona Beach Shores
ARRESTING OFFICER: MORIARTY, MATTHEW
ARRESTING OFFICER AGENCY: Daytona Beach Shores
LAST AGENCY INSPECTION: 11/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04216080A3
DRY GAS EXPIRY DATE: 03/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:33

SUBJECT'S LAST NAME: LEWIS

DATE/TIME OF TEST: 12/03/2017 04:06:35

Diagnostics Check	OK	03:55	
Air Blank	0.000	03:55	
Control Test	0.078	03:55	
Air Blank	0.000	03:56	
Subject Sample	# 1	03:59	No Sample Provided
Breath Volume	0.000	03:59	
Air Blank	0.000	04:00	
Air Blank	0.000	04:02	
Subject Sample	# 2	04:05	No Sample Provided
Breath Volume	0.000	04:05	
Air Blank	0.000	04:05	
Control Test	0.077	04:05	
Air Blank	0.000	04:06	
Diagnostics Check	OK	04:06	

Data Download Date/Time: 12/17/2017 02:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001242

INSTRUMENT REGISTERED TO: DAYTONA BEACH SHORES
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORIARTY, MATTHEW, J
BREATH TEST OPERATOR AGENCY: Daytona Beach Shores
ARRESTING OFFICER: BROWN, KRISTY
ARRESTING OFFICER AGENCY: Daytona Beach Shores
LAST AGENCY INSPECTION: 12/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04216080A3
DRY GAS EXPIRY DATE: 03/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:17

SUBJECT'S LAST NAME: BROOKS

DATE/TIME OF TEST: 12/26/2017 03:51:25

Diagnostics Check		OK	03:37	
Air Blank		0.000	03:37	
Control Test		0.077	03:38	
Air Blank		0.000	03:38	
Subject Sample	# 1	0.145	03:41	Volume Not Met
Breath Volume		0.457	03:41	
Air Blank		0.000	03:42	
Air Blank		0.000	03:44	
Subject Sample	# 2	0.179	03:46	
Breath Volume		1.976	03:46	
Air Blank		0.000	03:47	
Air Blank		0.000	03:49	
Subject Sample	# 3	0.169	03:49	
Breath Volume		1.542	03:49	
Air Blank		0.000	03:50	
Control Test		0.075	03:50	
Air Blank		0.000	03:51	
Diagnostics Check		OK	03:51	

Data Download Date/Time: 02/03/2018 21:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001242

INSTRUMENT REGISTERED TO: DAYTONA BEACH SHORES
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASTELLANO,VINCENT,J
BREATH TEST OPERATOR AGENCY: Daytona Beach Shores
ARRESTING OFFICER: CASTELLANO,VINCENT
ARRESTING OFFICER AGENCY: Daytona Beach Shores
LAST AGENCY INSPECTION: 12/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04216080A3
DRY GAS EXPIRY DATE: 03/05/2018
VIOLATION CODE: DUI
ARREST TIME: 08:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:35

SUBJECT'S LAST NAME: FLESCH

DATE/TIME OF TEST: 12/30/2017 09:02:23

Diagnostics Check	OK	08:55
Air Blank	0.000	08:56
Control Test	0.077	08:56
Air Blank	0.000	08:57
Subject Sample # 1	0.136	08:57
Breath Volume	2.437	08:57
Air Blank	0.000	08:58
Air Blank	0.000	09:00
Subject Sample # 2	0.141	09:00
Breath Volume	1.468	09:00
Air Blank	0.000	09:01
Control Test	0.077	09:01
Air Blank	0.000	09:02
Diagnostics Check	OK	09:02

Data Download Date/Time: 02/03/2018 21:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001243

INSTRUMENT REGISTERED TO: EDGEWATER P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SLOAN,JULIAN,E
BREATH TEST OPERATOR AGENCY: Edgewater PD
ARRESTING OFFICER: SLOAN,JULIAN
ARRESTING OFFICER AGENCY: Edgewater PD
LAST AGENCY INSPECTION: 11/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 765462
DRY GAS EXPIRY DATE: 04/13/2020
VIOLATION CODE: DUI
ARREST TIME: 18:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:13

SUBJECT'S LAST NAME: WALTERS

DATE/TIME OF TEST: 12/20/2017 19:42:37

Diagnostics Check	OK	19:35
Air Blank	0.000	19:36
Control Test	0.079	19:36
Air Blank	0.000	19:36
Subject Sample # 1	0.237	19:37
Breath Volume	1.804	19:37
Air Blank	0.000	19:38
Air Blank	0.000	19:40
Subject Sample # 2	0.252	19:40
Breath Volume	1.808	19:40
Air Blank	0.000	19:41
Control Test	0.076	19:42
Air Blank	0.000	19:42
Diagnostics Check	OK	19:42

Data Download Date/Time: 12/28/2017 23:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001243

INSTRUMENT REGISTERED TO: EDGEWATER P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BINZ,ANTHONY,E
BREATH TEST OPERATOR AGENCY: Edgewater PD
ARRESTING OFFICER: POTTER,KYLE
ARRESTING OFFICER AGENCY: Edgewater PD
LAST AGENCY INSPECTION: 11/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 765462
DRY GAS EXPIRY DATE: 04/13/2020
VIOLATION CODE: DUI
ARREST TIME: 15:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:57

SUBJECT'S LAST NAME: GRONDIN

DATE/TIME OF TEST: 12/24/2017 16:23:53

Diagnostics Check	OK	16:17
Air Blank	0.000	16:17
Control Test	0.079	16:18
Air Blank	0.000	16:18
Subject Sample # 1	0.084	16:19
Breath Volume	2.957	16:19
Air Blank	0.000	16:20
Air Blank	0.000	16:21
Subject Sample # 2	0.084	16:22
Breath Volume	2.261	16:22
Air Blank	0.000	16:22
Control Test	0.079	16:23
Air Blank	0.000	16:23
Diagnostics Check	OK	16:23

Data Download Date/Time: 12/28/2017 23:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001243

INSTRUMENT REGISTERED TO: EDGEWATER P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SLOAN,JULIAN,E
BREATH TEST OPERATOR AGENCY: Edgewater PD
ARRESTING OFFICER: WRIGHT,CORDARREL
ARRESTING OFFICER AGENCY: Edgewater PD
LAST AGENCY INSPECTION: 12/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 765462
DRY GAS EXPIRY DATE: 04/13/2020
VIOLATION CODE: DUI
ARREST TIME: 19:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:57

SUBJECT'S LAST NAME: TOTH

DATE/TIME OF TEST: 12/30/2017 20:25:21

Diagnostics Check	OK	20:18
Air Blank	0.000	20:19
Control Test	0.079	20:19
Air Blank	0.000	20:20
Subject Sample # 1	0.165	20:20
Breath Volume	1.253	20:20
Air Blank	0.000	20:21
Air Blank	0.000	20:23
Subject Sample # 2	0.169	20:23
Breath Volume	1.828	20:23
Air Blank	0.000	20:24
Control Test	0.079	20:24
Air Blank	0.000	20:25
Diagnostics Check	OK	20:25

Data Download Date/Time: 01/21/2018 10:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001245

INSTRUMENT REGISTERED TO: PORT ORANGE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BENEZETTE,BENJAMIN,J
BREATH TEST OPERATOR AGENCY: Port Orange PD
ARRESTING OFFICER: EDWARDS,LANDON
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 11/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 609878
DRY GAS EXPIRY DATE: 11/01/2019
VIOLATION CODE: DUI
ARREST TIME: 01:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: STARK

DATE/TIME OF TEST: 12/07/2017 02:40:45

Diagnostics Check	OK	02:34
Air Blank	0.000	02:34
Control Test	0.081	02:34
Air Blank	0.000	02:35
Subject Sample # 1	0.202	02:36
Breath Volume	1.312	02:36
Air Blank	0.000	02:36
Air Blank	0.000	02:38
Subject Sample # 2	0.211	02:39
Breath Volume	1.566	02:39
Air Blank	0.000	02:39
Control Test	0.079	02:40
Air Blank	0.000	02:40
Diagnostics Check	OK	02:40

Data Download Date/Time: 12/29/2017 18:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001245

INSTRUMENT REGISTERED TO: PORT ORANGE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NOLAN,JONATHAN,W
BREATH TEST OPERATOR AGENCY: Port Orange PD
ARRESTING OFFICER: RAYNOR,JASON
ARRESTING OFFICER AGENCY: Port Orange PD
LAST AGENCY INSPECTION: 11/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 609878
DRY GAS EXPIRY DATE: 11/01/2019
VIOLATION CODE: DUI
ARREST TIME: 23:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:42

SUBJECT'S LAST NAME: BRASSE

DATE/TIME OF TEST: 12/23/2017 00:09:03

Diagnostics Check		OK	00:02
Air Blank		0.000	00:02
Control Test		0.081	00:03
Air Blank		0.000	00:03
Subject Sample	# 1	0.066	00:04
Breath Volume		1.617	00:04
Air Blank		0.000	00:05
Air Blank		0.000	00:06
Subject Sample	# 2	0.061	00:07
Breath Volume		1.285	00:07
Air Blank		0.000	00:08
Control Test		0.082	00:08
Air Blank		0.000	00:08
Diagnostics Check		OK	00:09

Data Download Date/Time: 12/29/2017 18:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001245

INSTRUMENT REGISTERED TO: PORT ORANGE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NOLAN,JONATHAN,W
BREATH TEST OPERATOR AGENCY: Port Orange PD
ARRESTING OFFICER: ALEVISATOS,JARRAD
ARRESTING OFFICER AGENCY: Port Orange PD
LAST AGENCY INSPECTION: 11/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 609878
DRY GAS EXPIRY DATE: 11/01/2019
VIOLATION CODE: DUI
ARREST TIME: 01:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:03

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 12/23/2017 02:35:44

Diagnostics Check		OK	02:24	
Air Blank		0.000	02:24	
Control Test		0.081	02:24	
Air Blank		0.000	02:25	
Subject Sample	# 1	0.176	02:28	Volume Not Met
Breath Volume		0.046	02:28	
Air Blank		0.000	02:29	
Air Blank		0.000	02:31	
Subject Sample	# 2	0.189	02:34	Volume Not Met
Breath Volume		0.691	02:34	
Air Blank		0.000	02:34	
Control Test		0.081	02:35	
Air Blank		0.000	02:35	
Diagnostics Check		OK	02:35	

Data Download Date/Time: 12/29/2017 18:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001245

INSTRUMENT REGISTERED TO: PORT ORANGE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BENEZETTE,BENJAMIN,J
BREATH TEST OPERATOR AGENCY: Port Orange PD
ARRESTING OFFICER: ROBINSON,IVAN
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 11/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 609878
DRY GAS EXPIRY DATE: 11/01/2019
VIOLATION CODE: DUI
ARREST TIME: 20:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:43

SUBJECT'S LAST NAME: MATTESON

DATE/TIME OF TEST: 12/26/2017 22:13:31

Diagnostics Check	OK	22:07
Air Blank	0.000	22:07
Control Test	0.080	22:07
Air Blank	0.000	22:08
Subject Sample # 1	0.185	22:08
Breath Volume	1.714	22:08
Air Blank	0.000	22:09
Air Blank	0.000	22:11
Subject Sample # 2	0.188	22:12
Breath Volume	1.808	22:12
Air Blank	0.000	22:12
Control Test	0.081	22:12
Air Blank	0.000	22:13
Diagnostics Check	OK	22:13

Data Download Date/Time: 12/29/2017 18:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001245

INSTRUMENT REGISTERED TO: PORT ORANGE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRISON, WILLIAM, C
BREATH TEST OPERATOR AGENCY: Port Orange PD
ARRESTING OFFICER: N/A, N/A
ARRESTING OFFICER AGENCY: Port Orange PD
LAST AGENCY INSPECTION: 11/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 609878
DRY GAS EXPIRY DATE: 11/01/2019
VIOLATION CODE: ADMIN
ARREST TIME: 12:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:20

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 12/29/2017 11:51:54

Diagnostics Check	OK	11:46
Air Blank	0.000	11:46
Control Test	0.082	11:46
Air Blank	0.000	11:47
Subject Sample # 1	0.000	11:47
Breath Volume	1.410	11:47
Air Blank	0.000	11:48
Air Blank	0.000	11:50
Subject Sample # 2	0.000	11:50
Breath Volume	1.839	11:50
Air Blank	0.000	11:51
Control Test	0.080	11:51
Air Blank	0.000	11:51
Diagnostics Check	OK	11:51

Data Download Date/Time: 12/29/2017 18:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001245

INSTRUMENT REGISTERED TO: PORT ORANGE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRISON, WILLIAM, C
BREATH TEST OPERATOR AGENCY: Port Orange PD
ARRESTING OFFICER: N/A, N/A
ARRESTING OFFICER AGENCY: Port Orange PD
LAST AGENCY INSPECTION: 11/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 609878
DRY GAS EXPIRY DATE: 11/01/2019
VIOLATION CODE: ADMIN
ARREST TIME: 12:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:30

SUBJECT'S LAST NAME: STARNES

DATE/TIME OF TEST: 12/29/2017 12:06:30

Diagnostics Check	OK	12:00
Air Blank	0.000	12:00
Control Test	0.081	12:00
Air Blank	0.000	12:01
Subject Sample # 1	0.000	12:02
Breath Volume	4.601	12:02
Air Blank	0.000	12:02
Air Blank	0.000	12:04
Subject Sample # 2	0.000	12:05
Breath Volume	2.078	12:05
Air Blank	0.000	12:05
Control Test	0.080	12:05
Air Blank	0.000	12:06
Diagnostics Check	OK	12:06

Data Download Date/Time: 12/29/2017 18:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001250

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINA,MARMILLY,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: TRABUE,THOMAS
ARRESTING OFFICER AGENCY: Pembroke Pines PD
LAST AGENCY INSPECTION: 11/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 572920
DRY GAS EXPIRY DATE: 09/13/2019
VIOLATION CODE: DUI
ARREST TIME: 05:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:30

SUBJECT'S LAST NAME: POTPROCKY

DATE/TIME OF TEST: 12/04/2017 07:02:38

Diagnostics Check	OK	06:55
Air Blank	0.000	06:56
Control Test	0.080	06:56
Air Blank	0.000	06:56
Subject Sample # 1	0.057	06:57
Breath Volume	1.750	06:57
Air Blank	0.000	06:58
Air Blank	0.000	07:00
Subject Sample # 2	0.062	07:01
Breath Volume	1.238	07:01
Air Blank	0.000	07:01
Control Test	0.081	07:02
Air Blank	0.000	07:02
Diagnostics Check	OK	07:02

Data Download Date/Time: 12/13/2017 12:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001250

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINA,MARMILLY,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SMITH,STANLEY
ARRESTING OFFICER AGENCY: Hallandale Beach PD
LAST AGENCY INSPECTION: 11/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 572920
DRY GAS EXPIRY DATE: 09/13/2019
VIOLATION CODE: DUI
ARREST TIME: 21:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 12/09/2017 22:17:24

Diagnostics Check	OK	22:11
Air Blank	0.000	22:11
Control Test	0.080	22:12
Air Blank	0.000	22:12
Subject Sample # 1	0.240	22:12
Breath Volume	2.089	22:12
Air Blank	0.000	22:13
Air Blank	0.000	22:15
Subject Sample # 2	0.238	22:15
Breath Volume	1.339	22:15
Air Blank	0.000	22:16
Control Test	0.080	22:16
Air Blank	0.000	22:17
Diagnostics Check	OK	22:17

Data Download Date/Time: 12/13/2017 12:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001250

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINA,MARMILLY,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: PACK,JEFFERY
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 11/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 572920
DRY GAS EXPIRY DATE: 09/13/2019
VIOLATION CODE: DUI
ARREST TIME: 01:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: PIERRE

DATE/TIME OF TEST: 12/11/2017 02:26:40

Diagnostics Check	OK	02:19
Air Blank	0.000	02:20
Control Test	0.080	02:20
Air Blank	0.000	02:21
Subject Sample # 1	0.159	02:21
Breath Volume	1.738	02:21
Air Blank	0.000	02:22
Air Blank	0.000	02:24
Subject Sample # 2	0.166	02:25
Breath Volume	1.546	02:25
Air Blank	0.000	02:25
Control Test	0.081	02:26
Air Blank	0.000	02:26
Diagnostics Check	OK	02:26

Data Download Date/Time: 12/13/2017 12:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001250

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: NAVARRO,CASSIMIRO
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 11/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 572920
DRY GAS EXPIRY DATE: 09/13/2019
VIOLATION CODE: DUI
ARREST TIME: 19:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:12

SUBJECT'S LAST NAME: PAYNE

DATE/TIME OF TEST: 12/11/2017 20:45:43

Diagnostics Check	OK	20:38
Air Blank	0.000	20:39
Control Test	0.081	20:39
Air Blank	0.000	20:40
Subject Sample # 1	0.185	20:41
Breath Volume	2.710	20:41
Air Blank	0.000	20:41
Air Blank	0.000	20:43
Subject Sample # 2	0.193	20:44
Breath Volume	2.628	20:44
Air Blank	0.000	20:44
Control Test	0.081	20:45
Air Blank	0.000	20:45
Diagnostics Check	OK	20:45

Data Download Date/Time: 12/13/2017 12:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001250

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARNANDAN,INDROWTIE,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: OLEA,ELIEZER
ARRESTING OFFICER AGENCY: Pembroke Pines PD
LAST AGENCY INSPECTION: 11/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 572920
DRY GAS EXPIRY DATE: 09/13/2019
VIOLATION CODE: DUI
ARREST TIME: 17:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:18

SUBJECT'S LAST NAME: ESTLOW

DATE/TIME OF TEST: 12/12/2017 18:44:48

Diagnostics Check	OK	18:38
Air Blank	0.000	18:38
Control Test	0.080	18:39
Air Blank	0.000	18:39
Subject Sample # 1	0.224	18:40
Breath Volume	2.871	18:40
Air Blank	0.000	18:40
Air Blank	0.000	18:42
Subject Sample # 2	0.226	18:43
Breath Volume	3.183	18:43
Air Blank	0.000	18:43
Control Test	0.084	18:44
Air Blank	0.000	18:44
Diagnostics Check	OK	18:44

Data Download Date/Time: 12/13/2017 12:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001250

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: MCGILLICUDDY,STEVEN
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 11/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 572920
DRY GAS EXPIRY DATE: 09/13/2019
VIOLATION CODE: DUI
ARREST TIME: 04:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:34

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 12/13/2017 06:01:57

Diagnostics Check	OK	05:55
Air Blank	0.000	05:55
Control Test	0.080	05:56
Air Blank	0.000	05:56
Subject Sample # 1	0.262	05:57
Breath Volume	2.414	05:57
Air Blank	0.000	05:57
Air Blank	0.000	05:59
Subject Sample # 2	0.265	06:00
Breath Volume	3.050	06:00
Air Blank	0.000	06:01
Control Test	0.082	06:01
Air Blank	0.000	06:01
Diagnostics Check	OK	06:01

Data Download Date/Time: 12/13/2017 12:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001250

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: DEDEJ,MEDJON
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 11/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 572920
DRY GAS EXPIRY DATE: 09/13/2019
VIOLATION CODE: DUI
ARREST TIME: 07:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:10

SUBJECT'S LAST NAME: SOMA

DATE/TIME OF TEST: 12/13/2017 09:39:06

Diagnostics Check	OK	09:31
Air Blank	0.000	09:31
Control Test	0.081	09:32
Air Blank	0.000	09:32
Subject Sample # 1	0.000	09:34
Breath Volume	2.328	09:34
Air Blank	0.000	09:35
Air Blank	0.000	09:37
Subject Sample # 2	0.000	09:37
Breath Volume	1.980	09:37
Air Blank	0.000	09:38
Control Test	0.085	09:38
Air Blank	0.000	09:39
Diagnostics Check	OK	09:39

Data Download Date/Time: 12/13/2017 12:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001250

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: TRABUE,THOMAS
ARRESTING OFFICER AGENCY: Pembroke Pines PD
LAST AGENCY INSPECTION: 12/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 572920
DRY GAS EXPIRY DATE: 09/13/2019
VIOLATION CODE: DUI
ARREST TIME: 01:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:39

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 12/15/2017 03:03:46

Diagnostics Check	OK	03:00	
Air Blank	0.000	03:00	
Control Test	0.079	03:01	
Air Blank	0.000	03:01	
Subject Sample	# 1	03:02	Slope Not Met
Breath Volume	1.296	03:02	
Air Blank	0.000	03:02	
Control Test	0.079	03:03	
Air Blank	0.000	03:03	
Diagnostics Check	OK	03:03	

Data Download Date/Time: 01/10/2018 13:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001250

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: TRABUE,THOMAS
ARRESTING OFFICER AGENCY: Pembroke Pines PD
LAST AGENCY INSPECTION: 12/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 572920
DRY GAS EXPIRY DATE: 09/13/2019
VIOLATION CODE: DUI
ARREST TIME: 01:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:04

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 12/15/2017 03:36:05

Diagnostics Check	OK	03:29
Air Blank	0.000	03:30
Control Test	0.079	03:30
Air Blank	0.000	03:30
Subject Sample # 1	0.246	03:31
Breath Volume	3.308	03:31
Air Blank	0.000	03:32
Air Blank	0.000	03:33
Subject Sample # 2	0.253	03:34
Breath Volume	2.707	03:34
Air Blank	0.000	03:35
Control Test	0.081	03:35
Air Blank	0.000	03:36
Diagnostics Check	OK	03:36

Data Download Date/Time: 01/10/2018 13:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001250

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: NEWTON,PATRICK
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 12/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 572920
DRY GAS EXPIRY DATE: 09/13/2019
VIOLATION CODE: DUI
ARREST TIME: 03:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: CORMIER

DATE/TIME OF TEST: 12/16/2017 03:52:39

Diagnostics Check		OK	03:40	
Air Blank		0.000	03:41	
Control Test		0.080	03:41	
Air Blank		0.000	03:42	
Subject Sample	# 1	0.126	03:45	Volume Not Met
Breath Volume		0.125	03:45	
Air Blank		0.000	03:46	
Air Blank		0.000	03:47	
Subject Sample	# 2	0.132	03:50	Volume Not Met
Breath Volume		0.320	03:50	
Air Blank		0.000	03:51	
Control Test		0.085	03:52	
Air Blank		0.000	03:52	
Diagnostics Check		OK	03:52	

Data Download Date/Time: 01/10/2018 13:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001250

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: NEWTON,PATRICK
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 12/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 572920
DRY GAS EXPIRY DATE: 09/13/2019
VIOLATION CODE: DUI
ARREST TIME: 03:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: CORMIER

DATE/TIME OF TEST: 12/16/2017 04:07:03

Diagnostics Check		OK	03:55	
Air Blank		0.000	03:55	
Control Test		0.081	03:56	
Air Blank		0.000	03:56	
Subject Sample	# 1	0.123	03:59	Volume Not Met
Breath Volume		0.132	03:59	
Air Blank		0.000	04:00	
Air Blank		0.000	04:02	
Subject Sample	# 2	0.138	04:05	Volume Not Met
Breath Volume		0.789	04:05	
Air Blank		0.000	04:06	
Control Test		0.083	04:06	
Air Blank		0.000	04:06	
Diagnostics Check		OK	04:07	

Data Download Date/Time: 01/10/2018 13:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001250

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: DALTON,RYAN
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 12/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 572920
DRY GAS EXPIRY DATE: 09/13/2019
VIOLATION CODE: DUI
ARREST TIME: 03:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:35

SUBJECT'S LAST NAME: SPAULDING

DATE/TIME OF TEST: 12/17/2017 05:01:56

Diagnostics Check	OK	04:55
Air Blank	0.000	04:56
Control Test	0.080	04:56
Air Blank	0.000	04:56
Subject Sample # 1	0.218	04:57
Breath Volume	3.054	04:57
Air Blank	0.000	04:58
Air Blank	0.000	04:59
Subject Sample # 2	0.213	05:00
Breath Volume	3.125	05:00
Air Blank	0.000	05:01
Control Test	0.083	05:01
Air Blank	0.000	05:01
Diagnostics Check	OK	05:01

Data Download Date/Time: 01/10/2018 13:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001250

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ROJAS,RANDY
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 12/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 572920
DRY GAS EXPIRY DATE: 09/13/2019
VIOLATION CODE: DUI
ARREST TIME: 04:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:55

SUBJECT'S LAST NAME: QUINN

DATE/TIME OF TEST: 12/19/2017 05:23:18

Diagnostics Check	OK	05:16
Air Blank	0.000	05:16
Control Test	0.081	05:16
Air Blank	0.000	05:17
Subject Sample # 1	0.194	05:18
Breath Volume	1.992	05:18
Air Blank	0.000	05:19
Air Blank	0.000	05:21
Subject Sample # 2	0.199	05:21
Breath Volume	1.863	05:21
Air Blank	0.000	05:22
Control Test	0.085	05:22
Air Blank	0.000	05:23
Diagnostics Check	OK	05:23

Data Download Date/Time: 01/10/2018 13:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001250

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: DALTON,RYAN
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 12/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 572920
DRY GAS EXPIRY DATE: 09/13/2019
VIOLATION CODE: DUI
ARREST TIME: 04:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:41

SUBJECT'S LAST NAME: SANCHEZ-FRANCO

DATE/TIME OF TEST: 12/22/2017 05:08:57

Diagnostics Check	OK	05:01
Air Blank	0.000	05:02
Control Test	0.079	05:02
Air Blank	0.000	05:02
Subject Sample # 1	0.151	05:04
Breath Volume	2.238	05:04
Air Blank	0.000	05:04
Air Blank	0.000	05:06
Subject Sample # 2	0.144	05:07
Breath Volume	2.363	05:07
Air Blank	0.000	05:08
Control Test	0.084	05:08
Air Blank	0.000	05:08
Diagnostics Check	OK	05:08

Data Download Date/Time: 01/10/2018 13:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001250

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINA,MARMILLY,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SAPP,JOSHUA
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 12/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 572920
DRY GAS EXPIRY DATE: 09/13/2019
VIOLATION CODE: DUI
ARREST TIME: 19:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:25

SUBJECT'S LAST NAME: BRANCHEDOR

DATE/TIME OF TEST: 12/23/2017 21:01:17

Diagnostics Check		OK	20:49	
Air Blank		0.000	20:50	
Control Test		0.080	20:50	
Air Blank		0.000	20:50	
Subject Sample	# 1	0.228	20:53	Volume Not Met
Breath Volume		0.128	20:53	
Air Blank		0.000	20:54	
Air Blank		0.000	20:56	
Subject Sample	# 2	0.244	20:59	Volume Not Met
Breath Volume		0.082	20:59	
Air Blank		0.000	21:00	
Control Test		0.082	21:00	
Air Blank		0.000	21:01	
Diagnostics Check		OK	21:01	

Data Download Date/Time: 01/10/2018 13:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001250

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINA,MARMILLY,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SAPP,JOSHUA
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 12/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 572920
DRY GAS EXPIRY DATE: 09/13/2019
VIOLATION CODE: DUI
ARREST TIME: 19:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:25

SUBJECT'S LAST NAME: BRANCHEDOR

DATE/TIME OF TEST: 12/23/2017 21:18:29

Diagnostics Check		OK	21:04	
Air Blank		0.000	21:05	
Control Test		0.079	21:05	
Air Blank		0.000	21:06	
Subject Sample	# 1	0.234	21:09	Volume Not Met
Breath Volume		0.367	21:09	
Air Blank		0.000	21:10	
Air Blank		0.000	21:11	
Subject Sample	# 2	0.251	21:13	
Breath Volume		1.812	21:13	
Air Blank		0.000	21:14	
Air Blank		0.000	21:16	
Subject Sample	# 3	0.246	21:16	
Breath Volume		1.347	21:16	
Air Blank		0.000	21:17	
Control Test		0.081	21:17	
Air Blank		0.000	21:18	
Diagnostics Check		OK	21:18	

Data Download Date/Time: 01/10/2018 13:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001250

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: HARRIS,MILTON
ARRESTING OFFICER AGENCY: Pembroke Pines PD
LAST AGENCY INSPECTION: 12/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 572920
DRY GAS EXPIRY DATE: 09/13/2019
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: ABBATTISTA

DATE/TIME OF TEST: 12/24/2017 03:57:58

Diagnostics Check		OK	03:46	
Air Blank		0.000	03:46	
Control Test		0.080	03:47	
Air Blank		0.000	03:47	
Subject Sample	# 1	0.188	03:50	Volume Not Met
Breath Volume		0.066	03:50	
Air Blank		0.000	03:51	
Air Blank		0.000	03:53	
Subject Sample	# 2		03:56	No Sample Provided
Breath Volume		0.000	03:56	
Air Blank		0.000	03:57	
Control Test		0.082	03:57	
Air Blank		0.000	03:57	
Diagnostics Check		OK	03:57	

Data Download Date/Time: 01/10/2018 13:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001250

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINA,MARMILLY,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SACKMANN,ANDREW
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 12/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 572920
DRY GAS EXPIRY DATE: 09/13/2019
VIOLATION CODE: DUI
ARREST TIME: 16:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:25

SUBJECT'S LAST NAME: SAENZ

DATE/TIME OF TEST: 12/26/2017 17:52:39

Diagnostics Check		OK	17:45
Air Blank		0.000	17:46
Control Test		0.080	17:46
Air Blank		0.000	17:46
Subject Sample	# 1	0.285	17:47
Breath Volume		1.464	17:47
Air Blank		0.000	17:48
Air Blank		0.000	17:50
Subject Sample	# 2	0.291	17:51
Breath Volume		1.851	17:51
Air Blank		0.000	17:51
Control Test		0.081	17:52
Air Blank		0.000	17:52
Diagnostics Check		OK	17:52

Data Download Date/Time: 01/10/2018 13:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001250

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SERPHOS,STEVEN,A
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: POPICK,SCOTT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 12/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 572920
DRY GAS EXPIRY DATE: 09/13/2019
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:11

SUBJECT'S LAST NAME: FLOWERS

DATE/TIME OF TEST: 12/27/2017 01:43:59

Diagnostics Check	OK	01:36	
Air Blank	0.000	01:36	
Control Test	0.080	01:36	
Air Blank	0.000	01:37	
Subject Sample	# 1 0.228	01:37	
Breath Volume	1.250	01:37	
Air Blank	0.000	01:38	
Air Blank	0.000	01:40	
Subject Sample	# 2 0.215	01:42	
Breath Volume	1.535	01:42	
Air Blank	0.000	01:43	
Control Test	0.086	01:43	Control Outside Tolerance
Air Blank	0.000	01:43	

Data Download Date/Time: 01/10/2018 13:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001250

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SERPHOS,STEVEN,A
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: POPICK,SCOTT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 12/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 572920
DRY GAS EXPIRY DATE: 09/13/2019
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:11

SUBJECT'S LAST NAME: FLOWERS

DATE/TIME OF TEST: 12/27/2017 01:52:16

Diagnostics Check	OK	01:45
Air Blank	0.000	01:46
Control Test	0.080	01:46
Air Blank	0.000	01:47
Subject Sample # 1	0.222	01:47
Breath Volume	1.105	01:47
Air Blank	0.000	01:48
Air Blank	0.000	01:50
Subject Sample # 2	0.223	01:50
Breath Volume	1.796	01:50
Air Blank	0.000	01:51
Control Test	0.081	01:51
Air Blank	0.000	01:52
Diagnostics Check	OK	01:52

Data Download Date/Time: 01/10/2018 13:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001250

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINA,MARMILLY,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SANTIAGO,CARLOS
ARRESTING OFFICER AGENCY: Pembroke Pines PD
LAST AGENCY INSPECTION: 12/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 572920
DRY GAS EXPIRY DATE: 09/13/2019
VIOLATION CODE: DUI
ARREST TIME: 19:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:40

SUBJECT'S LAST NAME: TORRES

DATE/TIME OF TEST: 12/29/2017 21:25:22

Diagnostics Check	OK	21:15	
Air Blank	0.000	21:16	
Control Test	0.081	21:16	
Air Blank	0.000	21:17	
Subject Sample	# 1	21:20	No Sample Provided
Breath Volume	0.000	21:20	
Air Blank	0.000	21:20	
Air Blank	0.000	21:22	
Subject Sample	# 2	21:24	Subject Test Refused
Breath Volume	0.000	21:24	
Air Blank	0.000	21:24	
Control Test	0.080	21:24	
Air Blank	0.000	21:25	
Diagnostics Check	OK	21:25	

Data Download Date/Time: 01/10/2018 13:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001250

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAPP,JOSHUA,R
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SAPP,JOSHUA
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 12/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 572920
DRY GAS EXPIRY DATE: 09/13/2019
VIOLATION CODE: DUI
ARREST TIME: 01:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: ARIAS-SCHREIBER WIES

DATE/TIME OF TEST: 12/30/2017 02:28:03

Diagnostics Check	OK	02:21
Air Blank	0.000	02:22
Control Test	0.080	02:22
Air Blank	0.000	02:22
Subject Sample # 1	0.025	02:23
Breath Volume	3.132	02:23
Air Blank	0.000	02:24
Air Blank	0.000	02:25
Subject Sample # 2	0.025	02:26
Breath Volume	2.058	02:26
Air Blank	0.000	02:27
Control Test	0.084	02:27
Air Blank	0.000	02:27
Diagnostics Check	OK	02:28

Data Download Date/Time: 01/10/2018 13:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001250

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINA,MARMILLY,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FITHIAN,JAY
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 12/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 572920
DRY GAS EXPIRY DATE: 09/13/2019
VIOLATION CODE: DUI
ARREST TIME: 03:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: RIOS

DATE/TIME OF TEST: 12/30/2017 04:31:36

Diagnostics Check	OK	04:25
Air Blank	0.000	04:25
Control Test	0.080	04:26
Air Blank	0.000	04:26
Subject Sample # 1	0.136	04:27
Breath Volume	2.125	04:27
Air Blank	0.000	04:27
Air Blank	0.000	04:29
Subject Sample # 2	0.132	04:30
Breath Volume	1.566	04:30
Air Blank	0.000	04:30
Control Test	0.080	04:31
Air Blank	0.000	04:31
Diagnostics Check	OK	04:31

Data Download Date/Time: 01/10/2018 13:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001250

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINA,MARMILLY,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: LLOYD,WAYNE
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 12/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 572920
DRY GAS EXPIRY DATE: 09/13/2019
VIOLATION CODE: DUI
ARREST TIME: 19:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:55

SUBJECT'S LAST NAME: PAUL

DATE/TIME OF TEST: 12/30/2017 20:33:22

Diagnostics Check	OK	20:26
Air Blank	0.000	20:27
Control Test	0.080	20:27
Air Blank	0.000	20:28
Subject Sample # 1	0.176	20:28
Breath Volume	1.628	20:28
Air Blank	0.000	20:29
Air Blank	0.000	20:31
Subject Sample # 2	0.181	20:31
Breath Volume	2.273	20:31
Air Blank	0.000	20:32
Control Test	0.080	20:32
Air Blank	0.000	20:33
Diagnostics Check	OK	20:33

Data Download Date/Time: 01/10/2018 13:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001253

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANNON,LAUREN,F
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: ROBIN,GASTON
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 11/16/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: TOMPKINS

DATE/TIME OF TEST: 12/07/2017 02:20:50

Diagnostics Check	OK	02:14
Air Blank	0.000	02:15
Control Test	0.079	02:15
Air Blank	0.000	02:15
Subject Sample # 1	0.224	02:16
Breath Volume	3.355	02:16
Air Blank	0.000	02:17
Air Blank	0.000	02:18
Subject Sample # 2	0.230	02:19
Breath Volume	3.484	02:19
Air Blank	0.000	02:19
Control Test	0.077	02:20
Air Blank	0.000	02:20
Diagnostics Check	OK	02:20

Data Download Date/Time: 01/02/2018 20:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001253

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWARD, WILLIAM,
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: HOWARD, WILLIAM
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 11/16/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: PAPCZYNSKI

DATE/TIME OF TEST: 12/16/2017 00:48:18

Diagnostics Check		OK	00:40
Air Blank		0.000	00:40
Control Test		0.079	00:41
Air Blank		0.000	00:41
Subject Sample	# 1	0.104	00:42
Breath Volume		2.695	00:42
Air Blank		0.000	00:42
Air Blank		0.000	00:44
Subject Sample	# 2	0.108	00:46
Breath Volume		3.019	00:46
Air Blank		0.000	00:47
Control Test		0.081	00:47
Air Blank		0.000	00:48
Diagnostics Check		OK	00:48

Data Download Date/Time: 01/02/2018 20:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001253

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWARD, WILLIAM,
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: HOWARD, WILLIAM
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 11/16/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 12/18/2017 03:45:57

Diagnostics Check	OK	03:39
Air Blank	0.000	03:39
Control Test	0.081	03:40
Air Blank	0.000	03:40
Subject Sample # 1	0.184	03:41
Breath Volume	1.976	03:41
Air Blank	0.000	03:42
Air Blank	0.000	03:43
Subject Sample # 2	0.178	03:44
Breath Volume	1.980	03:44
Air Blank	0.000	03:45
Control Test	0.080	03:45
Air Blank	0.000	03:45
Diagnostics Check	OK	03:45

Data Download Date/Time: 01/02/2018 20:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001253

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWARD, WILLIAM,
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: HOWARD, WILLIAM
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 11/16/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:42

SUBJECT'S LAST NAME: CASTRO

DATE/TIME OF TEST: 12/22/2017 04:10:41

Diagnostics Check	OK	04:03
Air Blank	0.000	04:03
Control Test	0.080	04:04
Air Blank	0.000	04:04
Subject Sample # 1	0.205	04:05
Breath Volume	2.375	04:05
Air Blank	0.000	04:06
Air Blank	0.000	04:08
Subject Sample # 2	0.200	04:09
Breath Volume	1.949	04:09
Air Blank	0.000	04:09
Control Test	0.079	04:10
Air Blank	0.000	04:10
Diagnostics Check	OK	04:10

Data Download Date/Time: 01/02/2018 20:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001253

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWARD, WILLIAM,
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: HOWARD, WILLIAM
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 12/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:59

SUBJECT'S LAST NAME: FIGUEROA

DATE/TIME OF TEST: 12/30/2017 21:26:02

Diagnostics Check	OK	21:19
Air Blank	0.000	21:19
Control Test	0.080	21:20
Air Blank	0.000	21:20
Subject Sample # 1	0.138	21:21
Breath Volume	2.062	21:21
Air Blank	0.000	21:22
Air Blank	0.000	21:24
Subject Sample # 2	0.137	21:24
Breath Volume	1.125	21:24
Air Blank	0.000	21:25
Control Test	0.081	21:25
Air Blank	0.000	21:25
Diagnostics Check	OK	21:26

Data Download Date/Time: 01/02/2018 20:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001254

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITEMAN,NICHOLIS,J
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: OCONNELL,SHANNON
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 11/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 23:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: TRACEY

DATE/TIME OF TEST: 12/02/2017 01:01:45

Diagnostics Check		OK	00:55
Air Blank		0.000	00:56
Control Test		0.080	00:56
Air Blank		0.000	00:56
Subject Sample	# 1	0.134	00:57
Breath Volume		1.433	00:57
Air Blank		0.000	00:57
Air Blank		0.000	00:59
Subject Sample	# 2	0.136	01:00
Breath Volume		1.777	01:00
Air Blank		0.000	01:00
Control Test		0.081	01:01
Air Blank		0.000	01:01
Diagnostics Check		OK	01:01

Data Download Date/Time: 12/21/2017 11:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001254

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITEMAN,NICHOLIS,J
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: SWOGGER,EDWARD
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 11/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 23:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:17

SUBJECT'S LAST NAME: SCHMINKE

DATE/TIME OF TEST: 12/12/2017 00:39:56

Diagnostics Check	OK	00:37	
Air Blank	0.000	00:37	
Control Test	0.082	00:37	
Air Blank	0.000	00:38	
Subject Sample	# 1	00:38	Subject Test Refused
Breath Volume	0.000	00:38	
Air Blank	0.000	00:39	
Control Test	0.081	00:39	
Air Blank	0.000	00:39	
Diagnostics Check	OK	00:39	

Data Download Date/Time: 12/21/2017 11:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001254

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUTIERREZ,KRISTINA,L
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: CONNELL,SHANNON
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 11/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 18:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:10

SUBJECT'S LAST NAME: CARR

DATE/TIME OF TEST: 12/17/2017 19:37:22

Diagnostics Check	OK	19:30
Air Blank	0.000	19:31
Control Test	0.081	19:31
Air Blank	0.000	19:32
Subject Sample # 1	0.151	19:32
Breath Volume	1.398	19:32
Air Blank	0.000	19:33
Air Blank	0.000	19:35
Subject Sample # 2	0.162	19:35
Breath Volume	1.402	19:35
Air Blank	0.000	19:36
Control Test	0.080	19:36
Air Blank	0.000	19:37
Diagnostics Check	OK	19:37

Data Download Date/Time: 12/21/2017 11:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELBON,NICHOLAS,C
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: AZEVEDO,FABIO
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 12/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:36

SUBJECT'S LAST NAME: NEWBERRY

DATE/TIME OF TEST: 12/23/2017 03:06:42

Diagnostics Check	OK	03:00
Air Blank	0.000	03:01
Control Test	0.081	03:01
Air Blank	0.000	03:01
Subject Sample # 1	0.161	03:02
Breath Volume	2.777	03:02
Air Blank	0.000	03:02
Air Blank	0.000	03:04
Subject Sample # 2	0.167	03:05
Breath Volume	3.707	03:05
Air Blank	0.000	03:05
Control Test	0.079	03:06
Air Blank	0.000	03:06
Diagnostics Check	OK	03:06

Data Download Date/Time: 12/28/2017 16:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,CALEB,N
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ELLER,ASHLEY
ARRESTING OFFICER AGENCY: Apopka PD
LAST AGENCY INSPECTION: 12/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:48

SUBJECT'S LAST NAME: DOMINGUEZGOMEZ

DATE/TIME OF TEST: 12/30/2017 00:19:13

Diagnostics Check	OK	00:12
Air Blank	0.000	00:13
Control Test	0.081	00:13
Air Blank	0.000	00:13
Subject Sample # 1	0.134	00:14
Breath Volume	2.683	00:14
Air Blank	0.000	00:15
Air Blank	0.000	00:17
Subject Sample # 2	0.131	00:17
Breath Volume	3.160	00:17
Air Blank	0.000	00:18
Control Test	0.080	00:18
Air Blank	0.000	00:19
Diagnostics Check	OK	00:19

Data Download Date/Time: 01/04/2018 16:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TIAMSON,DAWN,P
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HAWKINS,DEBRA
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 12/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:38

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 12/30/2017 03:14:13

Diagnostics Check	OK	03:02	
Air Blank	0.000	03:03	
Control Test	0.080	03:03	
Air Blank	0.000	03:04	
Subject Sample	# 1	03:07	No Sample Provided
Breath Volume	0.000	03:07	
Air Blank	0.000	03:07	
Air Blank	0.000	03:09	
Subject Sample	# 2	03:12	No Sample Provided
Breath Volume	0.000	03:12	
Air Blank	0.000	03:13	
Control Test	0.080	03:13	
Air Blank	0.000	03:14	
Diagnostics Check	OK	03:14	

Data Download Date/Time: 01/04/2018 16:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELBON,NICHOLAS,C
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: AZEVEDO,FABIO
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 11/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:34

SUBJECT'S LAST NAME: AHBIZ

DATE/TIME OF TEST: 12/02/2017 05:12:20

Diagnostics Check	OK	05:06
Air Blank	0.000	05:06
Control Test	0.080	05:07
Air Blank	0.000	05:07
Subject Sample # 1	0.112	05:07
Breath Volume	1.718	05:07
Air Blank	0.000	05:08
Air Blank	0.000	05:10
Subject Sample # 2	0.105	05:10
Breath Volume	2.203	05:10
Air Blank	0.000	05:11
Control Test	0.079	05:11
Air Blank	0.000	05:12
Diagnostics Check	OK	05:12

Data Download Date/Time: 12/07/2017 13:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,CALEB,N
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CARTER,ELWOOD
ARRESTING OFFICER AGENCY: Winter Garden PD
LAST AGENCY INSPECTION: 11/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:38

SUBJECT'S LAST NAME: HALL

DATE/TIME OF TEST: 12/03/2017 02:07:51

Diagnostics Check	OK	02:01
Air Blank	0.000	02:01
Control Test	0.079	02:02
Air Blank	0.000	02:02
Subject Sample # 1	0.154	02:03
Breath Volume	2.367	02:03
Air Blank	0.000	02:03
Air Blank	0.000	02:05
Subject Sample # 2	0.158	02:06
Breath Volume	2.402	02:06
Air Blank	0.000	02:06
Control Test	0.079	02:07
Air Blank	0.000	02:07
Diagnostics Check	OK	02:07

Data Download Date/Time: 12/07/2017 13:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,CALEB,N
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: QUARTERMAN,TAHVANI
ARRESTING OFFICER AGENCY: Apopka PD
LAST AGENCY INSPECTION: 11/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:53

SUBJECT'S LAST NAME: VILLARROEL

DATE/TIME OF TEST: 12/07/2017 01:31:06

Diagnostics Check	OK	01:19
Air Blank	0.000	01:19
Control Test	0.079	01:19
Air Blank	0.000	01:20
Subject Sample # 1	0.118	01:21
Breath Volume	2.433	01:21
Air Blank	0.000	01:21
Air Blank	0.000	01:23
Subject Sample # 2	0.089	01:25
Breath Volume	1.187	01:25
Air Blank	0.000	01:25
Air Blank	0.000	01:27
Subject Sample # 3	0.115	01:29
Breath Volume	3.406	01:29
Air Blank	0.000	01:30
Control Test	0.079	01:30
Air Blank	0.000	01:31
Diagnostics Check	OK	01:31

Data Download Date/Time: 12/07/2017 13:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELBON,NICHOLAS,C
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CUTLER,WILLIAM
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 11/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:15

SUBJECT'S LAST NAME: ALBIZU

DATE/TIME OF TEST: 12/09/2017 04:50:56

Diagnostics Check	OK	04:44
Air Blank	0.000	04:45
Control Test	0.079	04:45
Air Blank	0.000	04:46
Subject Sample # 1	0.202	04:46
Breath Volume	1.921	04:46
Air Blank	0.000	04:47
Air Blank	0.000	04:48
Subject Sample # 2	0.198	04:49
Breath Volume	1.421	04:49
Air Blank	0.000	04:50
Control Test	0.079	04:50
Air Blank	0.000	04:50
Diagnostics Check	OK	04:50

Data Download Date/Time: 12/12/2017 11:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,CALEB,N
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DRACH,STEFAN
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 11/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:19

SUBJECT'S LAST NAME: MCLEOD

DATE/TIME OF TEST: 12/10/2017 00:50:16

Diagnostics Check	OK	00:43
Air Blank	0.000	00:43
Control Test	0.079	00:44
Air Blank	0.000	00:44
Subject Sample # 1	0.190	00:45
Breath Volume	1.398	00:45
Air Blank	0.000	00:46
Air Blank	0.000	00:48
Subject Sample # 2	0.184	00:48
Breath Volume	1.117	00:48
Air Blank	0.000	00:49
Control Test	0.078	00:49
Air Blank	0.000	00:50
Diagnostics Check	OK	00:50

Data Download Date/Time: 12/12/2017 11:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SMITH,JAMES
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 11/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 05:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:40

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 12/10/2017 14:20:20

Diagnostics Check	OK	14:14
Air Blank	0.000	14:14
Control Test	0.080	14:15
Air Blank	0.000	14:15
Subject Sample # 1	0.000	14:16
Breath Volume	2.679	14:16
Air Blank	0.000	14:16
Air Blank	0.000	14:18
Subject Sample # 2	0.000	14:18
Breath Volume	1.867	14:18
Air Blank	0.000	14:19
Control Test	0.080	14:19
Air Blank	0.000	14:20
Diagnostics Check	OK	14:20

Data Download Date/Time: 12/12/2017 11:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: GORDON,CURTIS
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 11/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:03

SUBJECT'S LAST NAME: DOMENA MORALES

DATE/TIME OF TEST: 12/11/2017 02:34:08

Diagnostics Check	OK	02:28
Air Blank	0.000	02:28
Control Test	0.081	02:28
Air Blank	0.000	02:29
Subject Sample # 1	0.112	02:29
Breath Volume	1.753	02:29
Air Blank	0.000	02:30
Air Blank	0.000	02:32
Subject Sample # 2	0.120	02:32
Breath Volume	1.125	02:32
Air Blank	0.000	02:33
Control Test	0.079	02:33
Air Blank	0.000	02:34
Diagnostics Check	OK	02:34

Data Download Date/Time: 12/12/2017 11:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO SANTIAGO, ELDALY, J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: COLLINS, MARVIN
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 12/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:09

SUBJECT'S LAST NAME: FLORES

DATE/TIME OF TEST: 12/17/2017 04:40:12

Diagnostics Check	OK	04:34
Air Blank	0.000	04:34
Control Test	0.078	04:34
Air Blank	0.000	04:35
Subject Sample # 1	0.190	04:35
Breath Volume	1.402	04:35
Air Blank	0.000	04:36
Air Blank	0.000	04:38
Subject Sample # 2	0.180	04:38
Breath Volume	1.148	04:38
Air Blank	0.000	04:39
Control Test	0.079	04:39
Air Blank	0.000	04:40
Diagnostics Check	OK	04:40

Data Download Date/Time: 12/21/2017 17:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO SANTIAGO,ELDALY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: LACENTRA,DONALD
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 12/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 05:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:33

SUBJECT'S LAST NAME: RAMOS OCASIO

DATE/TIME OF TEST: 12/17/2017 07:02:39

Diagnostics Check	OK	06:56
Air Blank	0.000	06:56
Control Test	0.080	06:57
Air Blank	0.000	06:57
Subject Sample # 1	0.127	06:58
Breath Volume	1.671	06:58
Air Blank	0.000	06:58
Air Blank	0.000	07:00
Subject Sample # 2	0.123	07:01
Breath Volume	1.429	07:01
Air Blank	0.000	07:01
Control Test	0.079	07:02
Air Blank	0.000	07:02
Diagnostics Check	OK	07:02

Data Download Date/Time: 12/21/2017 17:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO SANTIAGO, ELDALY, J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: TORRES, NICOLE
ARRESTING OFFICER AGENCY: Windermere PD
LAST AGENCY INSPECTION: 12/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:03

SUBJECT'S LAST NAME: CEZARIO RAMOS

DATE/TIME OF TEST: 12/18/2017 02:34:52

Diagnostics Check	OK	02:27
Air Blank	0.000	02:28
Control Test	0.079	02:28
Air Blank	0.000	02:29
Subject Sample # 1	0.179	02:29
Breath Volume	1.468	02:29
Air Blank	0.000	02:30
Air Blank	0.000	02:32
Subject Sample # 2	0.184	02:33
Breath Volume	1.183	02:33
Air Blank	0.000	02:33
Control Test	0.078	02:34
Air Blank	0.000	02:34
Diagnostics Check	OK	02:34

Data Download Date/Time: 12/21/2017 17:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO SANTIAGO, ELDALY, J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MENDOZA, ANGEL
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 12/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:12

SUBJECT'S LAST NAME: RIVERA MELECIO

DATE/TIME OF TEST: 12/18/2017 04:47:30

Diagnostics Check		OK	04:35	
Air Blank		0.000	04:36	
Control Test		0.079	04:36	
Air Blank		0.000	04:37	
Subject Sample	# 1	0.107	04:40	Volume Not Met
Breath Volume		0.015	04:40	
Air Blank		0.000	04:40	
Air Blank		0.000	04:42	
Subject Sample	# 2	0.122	04:45	Volume Not Met
Breath Volume		0.050	04:45	
Air Blank		0.000	04:46	
Control Test		0.078	04:46	
Air Blank		0.000	04:47	
Diagnostics Check		OK	04:47	

Data Download Date/Time: 12/21/2017 17:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO SANTIAGO, ELDALY, J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: GORDON, CURTIS
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 12/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:46

SUBJECT'S LAST NAME: NEU

DATE/TIME OF TEST: 12/20/2017 05:16:09

Diagnostics Check		OK	05:09
Air Blank		0.000	05:09
Control Test		0.079	05:10
Air Blank		0.000	05:10
Subject Sample	# 1	0.182	05:11
Breath Volume		2.156	05:11
Air Blank		0.000	05:12
Air Blank		0.000	05:14
Subject Sample	# 2	0.172	05:14
Breath Volume		2.089	05:14
Air Blank		0.000	05:15
Control Test		0.078	05:15
Air Blank		0.000	05:16
Diagnostics Check		OK	05:16

Data Download Date/Time: 12/21/2017 17:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO SANTIAGO, ELDALY, J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DRACH, STEFAN
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 12/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:42

SUBJECT'S LAST NAME: BELLOWS

DATE/TIME OF TEST: 12/24/2017 03:15:04

Diagnostics Check	OK	03:07
Air Blank	0.000	03:08
Control Test	0.080	03:08
Air Blank	0.000	03:09
Subject Sample # 1	0.147	03:10
Breath Volume	2.636	03:10
Air Blank	0.000	03:10
Air Blank	0.000	03:12
Subject Sample # 2	0.143	03:13
Breath Volume	2.476	03:13
Air Blank	0.000	03:14
Control Test	0.080	03:14
Air Blank	0.000	03:14
Diagnostics Check	OK	03:15

Data Download Date/Time: 12/28/2017 16:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO SANTIAGO,ELDALY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: AZEVEDO,FABIO
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 12/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 05:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:19

SUBJECT'S LAST NAME: SAINPAULIN

DATE/TIME OF TEST: 12/24/2017 05:59:33

Diagnostics Check	OK	05:52
Air Blank	0.000	05:53
Control Test	0.080	05:53
Air Blank	0.000	05:53
Subject Sample # 1	0.160	05:54
Breath Volume	1.980	05:54
Air Blank	0.000	05:55
Air Blank	0.000	05:56
Subject Sample # 2	0.157	05:58
Breath Volume	2.011	05:58
Air Blank	0.000	05:58
Control Test	0.079	05:58
Air Blank	0.000	05:59
Diagnostics Check	OK	05:59

Data Download Date/Time: 12/28/2017 16:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO SANTIAGO, ELDALY, J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HAWKINS, DEBORAH
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 12/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 06:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:00

SUBJECT'S LAST NAME: ALBINO GONZALEZ

DATE/TIME OF TEST: 12/24/2017 07:32:57

Diagnostics Check	OK	07:25
Air Blank	0.000	07:26
Control Test	0.080	07:26
Air Blank	0.000	07:27
Subject Sample # 1	0.183	07:28
Breath Volume	2.941	07:28
Air Blank	0.000	07:29
Air Blank	0.000	07:30
Subject Sample # 2	0.165	07:31
Breath Volume	1.101	07:31
Air Blank	0.000	07:32
Control Test	0.079	07:32
Air Blank	0.000	07:32
Diagnostics Check	OK	07:32

Data Download Date/Time: 12/28/2017 16:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROARIO SANTIAGO,ELDALY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: LEROY,SANTIA
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 12/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: LEWIS

DATE/TIME OF TEST: 12/31/2017 03:48:05

Diagnostics Check	OK	03:40
Air Blank	0.000	03:41
Control Test	0.079	03:41
Air Blank	0.000	03:42
Subject Sample # 1	0.176	03:42
Breath Volume	1.128	03:42
Air Blank	0.000	03:43
Air Blank	0.000	03:44
Subject Sample # 2	0.184	03:46
Breath Volume	1.871	03:46
Air Blank	0.000	03:47
Control Test	0.079	03:47
Air Blank	0.000	03:48
Diagnostics Check	OK	03:48

Data Download Date/Time: 01/04/2018 16:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELBON,NICHOLAS,C
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ROBBINS,NATHAN
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 11/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:59

SUBJECT'S LAST NAME: MASTERS

DATE/TIME OF TEST: 12/01/2017 02:29:48

Diagnostics Check	OK	02:23
Air Blank	0.000	02:23
Control Test	0.080	02:23
Air Blank	0.000	02:24
Subject Sample # 1	0.172	02:24
Breath Volume	3.324	02:24
Air Blank	0.000	02:25
Air Blank	0.000	02:27
Subject Sample # 2	0.176	02:28
Breath Volume	4.589	02:28
Air Blank	0.000	02:28
Control Test	0.079	02:29
Air Blank	0.000	02:29
Diagnostics Check	OK	02:29

Data Download Date/Time: 12/07/2017 13:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OKEEFE,ALEXIS,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HALLOCK,S
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 11/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 12/01/2017 02:58:38

Diagnostics Check	OK	02:52
Air Blank	0.000	02:53
Control Test	0.079	02:53
Air Blank	0.000	02:53
Subject Sample # 1	0.120	02:54
Breath Volume	2.503	02:54
Air Blank	0.000	02:54
Air Blank	0.000	02:56
Subject Sample # 2	0.125	02:57
Breath Volume	2.671	02:57
Air Blank	0.000	02:57
Control Test	0.079	02:58
Air Blank	0.000	02:58
Diagnostics Check	OK	02:58

Data Download Date/Time: 12/07/2017 13:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,CALEB,N
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: AZEVEDO,FABIO
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 11/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:34

SUBJECT'S LAST NAME: CHRISTENSEN

DATE/TIME OF TEST: 12/02/2017 01:10:37

Diagnostics Check	OK	01:02
Air Blank	0.000	01:02
Control Test	0.080	01:02
Air Blank	0.000	01:03
Subject Sample # 1	0.156	01:06
Breath Volume	2.121	01:06
Air Blank	0.000	01:06
Air Blank	0.000	01:08
Subject Sample # 2	0.165	01:09
Breath Volume	2.933	01:09
Air Blank	0.000	01:09
Control Test	0.080	01:10
Air Blank	0.000	01:10
Diagnostics Check	OK	01:10

Data Download Date/Time: 12/07/2017 13:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ, MICHAEL, J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: RIVERA, LUIS
ARRESTING OFFICER AGENCY: Central FL Univ PD
LAST AGENCY INSPECTION: 11/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 07:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:20

SUBJECT'S LAST NAME: PENA MARQUEZ

DATE/TIME OF TEST: 12/02/2017 08:53:16

Diagnostics Check	OK	08:47
Air Blank	0.000	08:47
Control Test	0.080	08:47
Air Blank	0.000	08:48
Subject Sample # 1	0.124	08:48
Breath Volume	2.691	08:48
Air Blank	0.000	08:49
Air Blank	0.000	08:51
Subject Sample # 2	0.124	08:51
Breath Volume	2.632	08:51
Air Blank	0.000	08:52
Control Test	0.079	08:52
Air Blank	0.000	08:53
Diagnostics Check	OK	08:53

Data Download Date/Time: 12/07/2017 13:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ, MICHAEL, J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BAUER, TYLER
ARRESTING OFFICER AGENCY: Apopka PD
LAST AGENCY INSPECTION: 11/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 15:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:10

SUBJECT'S LAST NAME: GALLAGHER

DATE/TIME OF TEST: 12/02/2017 17:43:00

Diagnostics Check		OK	17:36
Air Blank		0.000	17:36
Control Test		0.080	17:36
Air Blank		0.000	17:37
Subject Sample	# 1	0.258	17:38
Breath Volume		1.542	17:38
Air Blank		0.000	17:39
Air Blank		0.000	17:40
Subject Sample	# 2	0.270	17:41
Breath Volume		1.519	17:41
Air Blank		0.000	17:42
Control Test		0.078	17:42
Air Blank		0.000	17:42
Diagnostics Check		OK	17:43

Data Download Date/Time: 12/07/2017 13:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: AZEVEDO,FABIO
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 11/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:47

SUBJECT'S LAST NAME: KESERICH

DATE/TIME OF TEST: 12/02/2017 22:17:33

Diagnostics Check	OK	22:11
Air Blank	0.000	22:11
Control Test	0.079	22:12
Air Blank	0.000	22:12
Subject Sample # 1	0.263	22:13
Breath Volume	1.984	22:13
Air Blank	0.000	22:13
Air Blank	0.000	22:15
Subject Sample # 2	0.265	22:15
Breath Volume	3.042	22:15
Air Blank	0.000	22:16
Control Test	0.080	22:16
Air Blank	0.000	22:17
Diagnostics Check	OK	22:17

Data Download Date/Time: 12/07/2017 13:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MCSPADDEN,JASON
ARRESTING OFFICER AGENCY: Winter Garden PD
LAST AGENCY INSPECTION: 11/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:05

SUBJECT'S LAST NAME: CHEBOTAR

DATE/TIME OF TEST: 12/03/2017 20:35:00

Diagnostics Check	OK	20:28
Air Blank	0.000	20:29
Control Test	0.080	20:29
Air Blank	0.000	20:30
Subject Sample # 1	0.174	20:30
Breath Volume	3.488	20:30
Air Blank	0.000	20:31
Air Blank	0.000	20:33
Subject Sample # 2	0.155	20:33
Breath Volume	1.726	20:33
Air Blank	0.000	20:34
Control Test	0.079	20:34
Air Blank	0.000	20:34
Diagnostics Check	OK	20:35

Data Download Date/Time: 12/07/2017 13:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ROTARIUS,JOSHUA
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 11/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:46

SUBJECT'S LAST NAME: NEWMAN

DATE/TIME OF TEST: 12/04/2017 02:15:41

Diagnostics Check		OK	02:09
Air Blank		0.000	02:10
Control Test		0.080	02:10
Air Blank		0.000	02:10
Subject Sample	# 1	0.203	02:11
Breath Volume		2.453	02:11
Air Blank		0.000	02:11
Air Blank		0.000	02:13
Subject Sample	# 2	0.202	02:14
Breath Volume		1.773	02:14
Air Blank		0.000	02:14
Control Test		0.080	02:15
Air Blank		0.000	02:15
Diagnostics Check		OK	02:15

Data Download Date/Time: 12/07/2017 13:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ, MICHAEL, J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: WONG SHUE, ANTHONY
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 11/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 07:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:49

SUBJECT'S LAST NAME: FIGUEROA

DATE/TIME OF TEST: 12/04/2017 08:20:01

Diagnostics Check	OK	08:14	
Air Blank	0.000	08:15	
Control Test	0.079	08:15	
Air Blank	0.000	08:16	
Subject Sample	# 1	08:18	Subject Test Refused
Breath Volume	0.000	08:18	
Air Blank	0.000	08:19	
Control Test	0.080	08:19	
Air Blank	0.000	08:19	
Diagnostics Check	OK	08:20	

Data Download Date/Time: 12/07/2017 13:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CANER,OSVALDO,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: COOP,JOHN
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 11/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:38

SUBJECT'S LAST NAME: PERSAUD

DATE/TIME OF TEST: 12/05/2017 04:09:34

Diagnostics Check	OK	04:03
Air Blank	0.000	04:03
Control Test	0.081	04:03
Air Blank	0.000	04:04
Subject Sample # 1	0.201	04:05
Breath Volume	1.542	04:05
Air Blank	0.000	04:05
Air Blank	0.000	04:07
Subject Sample # 2	0.204	04:08
Breath Volume	1.589	04:08
Air Blank	0.000	04:08
Control Test	0.080	04:08
Air Blank	0.000	04:09
Diagnostics Check	OK	04:09

Data Download Date/Time: 12/07/2017 13:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLS,ARMAND,G
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: NEWELL,JEFF
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 11/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 16:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:24

SUBJECT'S LAST NAME: CUPP

DATE/TIME OF TEST: 12/05/2017 17:58:46

Diagnostics Check	OK	17:49
Air Blank	0.000	17:49
Control Test	0.080	17:50
Air Blank	0.000	17:50
Subject Sample # 1	0.241	17:51
Breath Volume	1.976	17:51
Air Blank	0.000	17:52
Air Blank	0.000	17:53
Subject Sample # 2	0.208	17:54
Breath Volume	1.843	17:54
Air Blank	0.000	17:54
Air Blank	0.000	17:56
Subject Sample # 3	0.233	17:57
Breath Volume	3.007	17:57
Air Blank	0.000	17:57
Control Test	0.078	17:58
Air Blank	0.000	17:58
Diagnostics Check	OK	17:58

Data Download Date/Time: 12/07/2017 13:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HAWKINS, DEBORAH
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 11/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:47

SUBJECT'S LAST NAME: CRUZ

DATE/TIME OF TEST: 12/07/2017 01:17:33

Diagnostics Check	OK	01:10
Air Blank	0.000	01:10
Control Test	0.080	01:10
Air Blank	0.000	01:11
Subject Sample # 1	0.199	01:12
Breath Volume	1.277	01:12
Air Blank	0.000	01:12
Air Blank	0.000	01:14
Subject Sample # 2	0.192	01:15
Breath Volume	1.390	01:15
Air Blank	0.000	01:16
Control Test	0.079	01:16
Air Blank	0.000	01:17
Diagnostics Check	OK	01:17

Data Download Date/Time: 12/07/2017 13:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OKEEFE,ALEXIS,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BODE,JOSHUA
ARRESTING OFFICER AGENCY: Ocoee PD
LAST AGENCY INSPECTION: 11/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: BUONAURO

DATE/TIME OF TEST: 12/08/2017 02:50:42

Diagnostics Check	OK	02:43
Air Blank	0.000	02:43
Control Test	0.081	02:44
Air Blank	0.000	02:44
Subject Sample # 1	0.167	02:45
Breath Volume	1.296	02:45
Air Blank	0.000	02:46
Air Blank	0.000	02:48
Subject Sample # 2	0.172	02:49
Breath Volume	1.707	02:49
Air Blank	0.000	02:49
Control Test	0.080	02:50
Air Blank	0.000	02:50
Diagnostics Check	OK	02:50

Data Download Date/Time: 12/12/2017 12:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: LAWRENCE,NICK
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 11/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:08

SUBJECT'S LAST NAME: GRAZIANO

DATE/TIME OF TEST: 12/08/2017 22:36:49

Diagnostics Check	OK	22:31
Air Blank	0.000	22:31
Control Test	0.080	22:31
Air Blank	0.000	22:32
Subject Sample # 1	0.000	22:32
Breath Volume	2.289	22:32
Air Blank	0.000	22:33
Air Blank	0.000	22:35
Subject Sample # 2	0.000	22:35
Breath Volume	2.007	22:35
Air Blank	0.000	22:35
Control Test	0.080	22:36
Air Blank	0.000	22:36
Diagnostics Check	OK	22:36

Data Download Date/Time: 12/12/2017 12:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,CALEB,N
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ELLER,JESSICA
ARRESTING OFFICER AGENCY: Winter Park PD
LAST AGENCY INSPECTION: 11/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: RIJO

DATE/TIME OF TEST: 12/09/2017 04:01:55

Diagnostics Check	OK	03:53
Air Blank	0.000	03:54
Control Test	0.080	03:54
Air Blank	0.000	03:55
Subject Sample # 1	0.199	03:57
Breath Volume	1.761	03:57
Air Blank	0.000	03:57
Air Blank	0.000	03:59
Subject Sample # 2	0.191	04:00
Breath Volume	1.863	04:00
Air Blank	0.000	04:01
Control Test	0.077	04:01
Air Blank	0.000	04:01
Diagnostics Check	OK	04:01

Data Download Date/Time: 12/12/2017 12:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OKEEFE,ALEXIS,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HAMILTON,D
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 11/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: HARTWELL

DATE/TIME OF TEST: 12/09/2017 04:25:01

Diagnostics Check	OK	04:18
Air Blank	0.000	04:18
Control Test	0.079	04:19
Air Blank	0.000	04:19
Subject Sample # 1	0.151	04:20
Breath Volume	2.058	04:20
Air Blank	0.000	04:21
Air Blank	0.000	04:22
Subject Sample # 2	0.155	04:23
Breath Volume	2.187	04:23
Air Blank	0.000	04:24
Control Test	0.078	04:24
Air Blank	0.000	04:24
Diagnostics Check	OK	04:25

Data Download Date/Time: 12/12/2017 12:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ, MICHAEL, J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HERNANDEZ, YASSELL
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 11/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 08:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:45

SUBJECT'S LAST NAME: MORGAN

DATE/TIME OF TEST: 12/09/2017 14:28:25

Diagnostics Check	OK	14:13	
Air Blank	0.000	14:14	
Control Test	0.078	14:14	
Air Blank	0.000	14:15	
Subject Sample # 1	0.052	14:15	
Breath Volume	4.210	14:15	
Air Blank	0.000	14:16	
Air Blank	0.000	14:18	
Subject Sample # 2		14:21	No Sample Provided
Breath Volume	0.000	14:21	
Air Blank	0.000	14:21	
Air Blank	0.000	14:23	
Subject Sample # 3		14:27	No Sample Provided
Breath Volume	0.000	14:27	
Air Blank	0.000	14:27	
Control Test	0.079	14:27	
Air Blank	0.000	14:28	
Diagnostics Check	OK	14:28	

Data Download Date/Time: 12/12/2017 12:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ, MICHAEL, J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HERNANDEZ, YASSELL
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 11/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 08:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:20

SUBJECT'S LAST NAME: MORGAN

DATE/TIME OF TEST: 12/09/2017 14:50:48

Diagnostics Check	OK	14:44
Air Blank	0.000	14:45
Control Test	0.078	14:45
Air Blank	0.000	14:45
Subject Sample # 1	0.041	14:46
Breath Volume	2.437	14:46
Air Blank	0.000	14:47
Air Blank	0.000	14:48
Subject Sample # 2	0.039	14:49
Breath Volume	2.027	14:49
Air Blank	0.000	14:49
Control Test	0.077	14:50
Air Blank	0.000	14:50
Diagnostics Check	OK	14:50

Data Download Date/Time: 12/12/2017 12:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HALLOCK,STEPHEN
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 11/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:42

SUBJECT'S LAST NAME: FLORES GARCIA

DATE/TIME OF TEST: 12/10/2017 00:14:46

Diagnostics Check		OK	00:07
Air Blank		0.000	00:08
Control Test		0.078	00:08
Air Blank		0.000	00:08
Subject Sample	# 1	0.280	00:09
Breath Volume		2.050	00:09
Air Blank		0.000	00:10
Air Blank		0.000	00:12
Subject Sample	# 2	0.274	00:13
Breath Volume		1.933	00:13
Air Blank		0.000	00:13
Control Test		0.076	00:14
Air Blank		0.000	00:14
Diagnostics Check		OK	00:14

Data Download Date/Time: 12/12/2017 12:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,COURTNEY,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: URIBE,SERGIO
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 11/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:14

SUBJECT'S LAST NAME: ESPINALGRANDEZ

DATE/TIME OF TEST: 12/10/2017 03:45:12

Diagnostics Check	OK	03:38	
Air Blank	0.000	03:38	
Control Test	0.075	03:39	
Air Blank	0.000	03:39	
Subject Sample # 1	0.207	03:40	
Breath Volume	1.746	03:40	
Air Blank	0.000	03:41	
Air Blank	0.000	03:43	
Subject Sample # 2	0.212	03:43	
Breath Volume	1.890	03:43	
Air Blank	0.000	03:44	
Control Test	0.070	03:44	Control Outside Tolerance
Air Blank	0.000	03:45	

Data Download Date/Time: 12/12/2017 12:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SMITH,JAMES
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 11/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 05:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:40

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 12/10/2017 14:07:57

Diagnostics Check	OK	14:06	
Air Blank	0.000	14:07	
Control Test	0.000	14:07	Control Outside Tolerance
Air Blank	0.000	14:07	

Data Download Date/Time: 12/12/2017 12:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001260

INSTRUMENT REGISTERED TO: COCOA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,NICHOLAS,A
BREATH TEST OPERATOR AGENCY: Cocoa PD
ARRESTING OFFICER: HALL,NICHOLAS
ARRESTING OFFICER AGENCY: Cocoa PD
LAST AGENCY INSPECTION: 11/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06616080A2
DRY GAS EXPIRY DATE: 04/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: SANGSTER

DATE/TIME OF TEST: 12/12/2017 00:27:00

Diagnostics Check		OK	00:21
Air Blank		0.000	00:21
Control Test		0.080	00:21
Air Blank		0.000	00:22
Subject Sample	# 1	0.166	00:22
Breath Volume		2.570	00:22
Air Blank		0.000	00:23
Air Blank		0.000	00:25
Subject Sample	# 2	0.172	00:25
Breath Volume		3.105	00:25
Air Blank		0.000	00:26
Control Test		0.080	00:26
Air Blank		0.000	00:26
Diagnostics Check		OK	00:27

Data Download Date/Time: 12/16/2017 20:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001260

INSTRUMENT REGISTERED TO: COCOA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANDIS,BENJAMIN,M
BREATH TEST OPERATOR AGENCY: Cocoa PD
ARRESTING OFFICER: WADE,GLENN
ARRESTING OFFICER AGENCY: Cocoa PD
LAST AGENCY INSPECTION: 12/16/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06616080A2
DRY GAS EXPIRY DATE: 04/05/2018
VIOLATION CODE: DUI
ARREST TIME: 17:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:01

SUBJECT'S LAST NAME: FUENTES FRANCISCO

DATE/TIME OF TEST: 12/18/2017 17:33:47

Diagnostics Check	OK	17:21		
Air Blank	0.000	17:22		
Control Test	0.080	17:22		
Air Blank	0.000	17:22		
Subject Sample	# 1	17:26	No Sample Provided	
Breath Volume	0.000	17:26		
Air Blank	0.000	17:26		
Air Blank	0.000	17:28		
Subject Sample	# 2	0.362	17:29	No .020 agreement
Breath Volume	2.742	17:29		
Air Blank	0.000	17:29		
Air Blank	0.000	17:31		
Subject Sample	# 3	0.339	17:32	No .020 agreement
Breath Volume	2.292	17:32		
Air Blank	0.000	17:32		
Control Test	0.078	17:33		
Air Blank	0.000	17:33		
Diagnostics Check	OK	17:33		

Data Download Date/Time: 01/16/2018 21:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001261

INSTRUMENT REGISTERED TO: ROCKLEDGE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOMEZ,IVETTE,
BREATH TEST OPERATOR AGENCY: Rockledge PD
ARRESTING OFFICER: SMITH,JOSEPH
ARRESTING OFFICER AGENCY: Rockledge PD
LAST AGENCY INSPECTION: 11/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 572920
DRY GAS EXPIRY DATE: 09/13/2019
VIOLATION CODE: DUI
ARREST TIME: 00:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:58

SUBJECT'S LAST NAME: EDWARDS

DATE/TIME OF TEST: 12/05/2017 01:40:26

Diagnostics Check	OK	01:28		
Air Blank	0.000	01:29		
Control Test	0.080	01:29		
Air Blank	0.000	01:30		
Subject Sample	# 1	01:33	No Sample Provided	
Breath Volume	0.000	01:33		
Air Blank	0.000	01:33		
Air Blank	0.000	01:35		
Subject Sample	# 2	0.210	01:38	Slope Not Level
Breath Volume	1.433	01:38		
Air Blank	0.000	01:39		
Control Test	0.071	01:39	Control Outside Tolerance	
Air Blank	0.000	01:40		

Data Download Date/Time: 12/26/2017 10:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001261

INSTRUMENT REGISTERED TO: ROCKLEDGE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOMEZ,IVETTE,
BREATH TEST OPERATOR AGENCY: Rockledge PD
ARRESTING OFFICER: SMITH,JOSEPH
ARRESTING OFFICER AGENCY: Rockledge PD
LAST AGENCY INSPECTION: 11/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 572920
DRY GAS EXPIRY DATE: 09/13/2019
VIOLATION CODE: DUI
ARREST TIME: 00:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:58

SUBJECT'S LAST NAME: EDWARDS

DATE/TIME OF TEST: 12/05/2017 01:47:56

Diagnostics Check	OK	01:43	
Air Blank	0.000	01:44	
Control Test	0.077	01:44	
Air Blank	0.000	01:45	
Subject Sample	# 1	01:46	Subject Test Refused
Breath Volume	1.019	01:46	
Air Blank	0.000	01:47	
Control Test	0.074	01:47	Control Outside Tolerance
Air Blank	0.000	01:47	

Data Download Date/Time: 12/26/2017 10:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCLURE, CHARLES, W
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: MCCLURE, CHARLES
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 11/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 712031
DRY GAS EXPIRY DATE: 02/21/2020
VIOLATION CODE: DUI
ARREST TIME: 21:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:13

SUBJECT'S LAST NAME: CAMPBELL

DATE/TIME OF TEST: 12/09/2017 21:41:27

Diagnostics Check	OK	21:35	
Air Blank	0.000	21:35	
Control Test	0.080	21:36	
Air Blank	0.000	21:36	
Subject Sample	# 1 0.196	21:39	Volume Not Met
Breath Volume	0.050	21:39	
Air Blank		21:40	Purge Fail
Air Blank		21:41	Purge Fail

Data Download Date/Time: 12/28/2017 16:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCLURE,CHARLES,W
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: MCCLURE,CHARLES
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 11/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 712031
DRY GAS EXPIRY DATE: 02/21/2020
VIOLATION CODE: DUI
ARREST TIME: 21:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:13

SUBJECT'S LAST NAME: CAMPBELL

DATE/TIME OF TEST: 12/09/2017 21:51:30

Diagnostics Check	OK	21:43
Air Blank	0.000	21:43
Control Test	0.080	21:44
Air Blank	0.000	21:44
Subject Sample # 1	0.211	21:46
Breath Volume	1.875	21:46
Air Blank	0.000	21:47
Air Blank	0.000	21:49
Subject Sample # 2	0.207	21:49
Breath Volume	1.183	21:49
Air Blank	0.000	21:50
Control Test	0.079	21:50
Air Blank	0.000	21:51
Diagnostics Check	OK	21:51

Data Download Date/Time: 12/28/2017 16:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCLURE,CHARLES,W
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: MCCLURE,CHARLES
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 11/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 712031
DRY GAS EXPIRY DATE: 02/21/2020
VIOLATION CODE: DUI
ARREST TIME: 23:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:33

SUBJECT'S LAST NAME: HAVICE

DATE/TIME OF TEST: 12/14/2017 00:05:26

Diagnostics Check	OK	23:57
Air Blank	0.000	23:57
Control Test	0.081	23:58
Air Blank	0.000	23:58
Subject Sample # 2	0.159	23:59
Breath Volume	1.457	23:59
Air Blank	0.000	00:00
Air Blank	0.000	00:02
Subject Sample # 1	0.156	00:03
Breath Volume	1.292	00:03
Air Blank	0.000	00:04
Control Test	0.082	00:04
Air Blank	0.000	00:05
Diagnostics Check	OK	00:05

Data Download Date/Time: 12/28/2017 16:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCRANN,SEAN,M
BREATH TEST OPERATOR AGENCY: Satellite Beach PD
ARRESTING OFFICER: MCCRANN,SEAN
ARRESTING OFFICER AGENCY: Satellite Beach PD
LAST AGENCY INSPECTION: 11/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 712031
DRY GAS EXPIRY DATE: 02/21/2020
VIOLATION CODE: DUI
ARREST TIME: 23:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: COCKRELL

DATE/TIME OF TEST: 12/16/2017 01:01:58

Diagnostics Check	OK	00:55
Air Blank	0.000	00:55
Control Test	0.081	00:56
Air Blank	0.000	00:56
Subject Sample # 1	0.179	00:57
Breath Volume	2.035	00:57
Air Blank	0.000	00:57
Air Blank	0.000	00:59
Subject Sample # 2	0.174	01:00
Breath Volume	1.812	01:00
Air Blank	0.000	01:01
Control Test	0.081	01:01
Air Blank	0.000	01:01
Diagnostics Check	OK	01:01

Data Download Date/Time: 12/28/2017 16:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARLSON,EDWIN,J
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: CARLSON,EDWIN
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 11/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 712031
DRY GAS EXPIRY DATE: 02/21/2020
VIOLATION CODE: DUI
ARREST TIME: 20:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:27

SUBJECT'S LAST NAME: MELROY

DATE/TIME OF TEST: 12/20/2017 22:21:23

Diagnostics Check	OK	22:15	
Air Blank	0.000	22:15	
Control Test	0.080	22:15	
Air Blank	0.000	22:16	
Subject Sample	# 1 0.269	22:16	
Breath Volume	2.910	22:16	
Air Blank	0.000	22:17	
Air Blank	0.000	22:19	
Subject Sample	# 2 0.264	22:19	
Breath Volume	3.230	22:19	
Air Blank	0.000	22:20	
Control Test	0.072	22:20	Control Outside Tolerance
Air Blank	0.000	22:21	

Data Download Date/Time: 12/28/2017 16:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARLSON,EDWIN,J
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: CARLSON,EDWIN
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 11/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 712031
DRY GAS EXPIRY DATE: 02/21/2020
VIOLATION CODE: DUI
ARREST TIME: 20:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:27

SUBJECT'S LAST NAME: MELROY

DATE/TIME OF TEST: 12/20/2017 22:48:16

Diagnostics Check		OK	22:41
Air Blank		0.000	22:41
Control Test		0.078	22:42
Air Blank		0.000	22:42
Subject Sample	# 1	0.256	22:43
Breath Volume		2.367	22:43
Air Blank		0.000	22:44
Air Blank		0.000	22:45
Subject Sample	# 2	0.255	22:46
Breath Volume		2.187	22:46
Air Blank		0.000	22:47
Control Test		0.079	22:47
Air Blank		0.000	22:48
Diagnostics Check		OK	22:48

Data Download Date/Time: 12/28/2017 16:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCLURE, CHARLES, W
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: WHITE, TAYLOR
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 11/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 712031
DRY GAS EXPIRY DATE: 02/21/2020
VIOLATION CODE: DUI
ARREST TIME: 02:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:57

SUBJECT'S LAST NAME: PATTON

DATE/TIME OF TEST: 12/23/2017 03:37:07

Diagnostics Check		OK	03:27
Air Blank		0.000	03:28
Control Test		0.080	03:28
Air Blank		0.000	03:28
Subject Sample	# 1	0.194	03:29
Breath Volume		1.480	03:29
Air Blank		0.000	03:29
Air Blank		0.000	03:31
Subject Sample	# 2	0.218	03:32
Breath Volume		3.128	03:32
Air Blank		0.000	03:33
Air Blank		0.000	03:34
Subject Sample	# 3	0.220	03:35
Breath Volume		2.906	03:35
Air Blank		0.000	03:36
Control Test		0.081	03:36
Air Blank		0.000	03:37
Diagnostics Check		OK	03:37

Data Download Date/Time: 12/28/2017 16:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCLURE, CHARLES, W
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: PEREZ, VICTOR
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 11/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 712031
DRY GAS EXPIRY DATE: 02/21/2020
VIOLATION CODE: DUI
ARREST TIME: 19:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:35

SUBJECT'S LAST NAME: ROIG

DATE/TIME OF TEST: 12/23/2017 20:03:36

Diagnostics Check	OK	19:57
Air Blank	0.000	19:57
Control Test	0.080	19:58
Air Blank	0.000	19:58
Subject Sample # 1	0.302	19:59
Breath Volume	2.859	19:59
Air Blank	0.000	19:59
Air Blank	0.000	20:01
Subject Sample # 2	0.297	20:02
Breath Volume	2.960	20:02
Air Blank	0.000	20:02
Control Test	0.080	20:02
Air Blank	0.000	20:03
Diagnostics Check	OK	20:03

Data Download Date/Time: 12/28/2017 16:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NOLES,GABRIEL,D
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: DUMMER,LUKE
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 12/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 712031
DRY GAS EXPIRY DATE: 02/21/2020
VIOLATION CODE: DUI
ARREST TIME: 04:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:39

SUBJECT'S LAST NAME: DIBENEDETTO

DATE/TIME OF TEST: 12/29/2017 05:13:34

Diagnostics Check	OK	05:07	
Air Blank	0.000	05:07	
Control Test	0.080	05:08	
Air Blank	0.000	05:08	
Subject Sample # 1	0.000	05:11	Volume Not Met
Breath Volume	0.167	05:11	
Air Blank		05:12	RFI Detect
Air Blank	0.000	05:12	
Control Test	0.084	05:12	
Air Blank	0.000	05:13	
Diagnostics Check	OK	05:13	

Data Download Date/Time: 01/30/2018 13:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NOLES,GABRIEL,D
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: DUMMER,LUKE
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 12/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 712031
DRY GAS EXPIRY DATE: 02/21/2020
VIOLATION CODE: DUI
ARREST TIME: 04:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:39

SUBJECT'S LAST NAME: DIBENEDETTO

DATE/TIME OF TEST: 12/29/2017 05:33:46

Diagnostics Check		OK	05:22	
Air Blank		0.000	05:22	
Control Test		0.080	05:23	
Air Blank		0.000	05:23	
Subject Sample	# 1	0.000	05:26	Volume Not Met
Breath Volume		0.289	05:26	
Air Blank		0.000	05:27	
Air Blank		0.000	05:29	
Subject Sample	# 2		05:32	No Sample Provided
Breath Volume		0.000	05:32	
Air Blank		0.000	05:32	
Control Test		0.080	05:33	
Air Blank		0.000	05:33	
Diagnostics Check		OK	05:33	

Data Download Date/Time: 01/30/2018 13:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARLSON, EDWIN, J
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: KOUBEK, STEPHEN
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 12/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 712031
DRY GAS EXPIRY DATE: 02/21/2020
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: COLLADO-LOPEZ

DATE/TIME OF TEST: 12/30/2017 01:04:50

Diagnostics Check		OK	00:55
Air Blank		0.000	00:55
Control Test		0.081	00:55
Air Blank		0.000	00:56
Subject Sample	# 1	0.216	00:58
Breath Volume		3.058	00:58
Air Blank		0.000	00:59
Air Blank		0.000	01:00
Subject Sample	# 2	0.198	01:03
Breath Volume		2.785	01:03
Air Blank		0.000	01:03
Control Test		0.079	01:04
Air Blank		0.000	01:04
Diagnostics Check		OK	01:04

Data Download Date/Time: 01/30/2018 13:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HESSION,CHRISTOPHER,S
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: HESSION,CHRISTOPHER
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 12/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 712031
DRY GAS EXPIRY DATE: 02/21/2020
VIOLATION CODE: DUI
ARREST TIME: 22:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: TRAVIS

DATE/TIME OF TEST: 12/30/2017 22:45:45

Diagnostics Check	OK	22:42	
Air Blank	0.000	22:42	
Control Test	0.080	22:43	
Air Blank	0.000	22:43	
Subject Sample	# 1	22:44	Subject Test Refused
Breath Volume	0.000	22:44	
Air Blank	0.000	22:44	
Control Test	0.080	22:45	
Air Blank	0.000	22:45	
Diagnostics Check	OK	22:45	

Data Download Date/Time: 01/30/2018 13:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001263

INSTRUMENT REGISTERED TO: INDIAN HARBOUR PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GEHRING,ALEXANDER,L
BREATH TEST OPERATOR AGENCY: Indian Harbour Beach
ARRESTING OFFICER: THOMPSON,ERIC
ARRESTING OFFICER AGENCY: Indian Harbour Beach
LAST AGENCY INSPECTION: 11/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:57

SUBJECT'S LAST NAME: GEHRING

DATE/TIME OF TEST: 12/08/2017 01:23:26

Diagnostics Check	OK	01:18	
Air Blank	0.000	01:18	
Control Test	0.081	01:18	
Air Blank	0.000	01:19	
Subject Sample	# 1	01:22	Subject Test Refused
Breath Volume	0.000	01:22	
Air Blank	0.000	01:22	
Control Test	0.081	01:22	
Air Blank	0.000	01:23	
Diagnostics Check	OK	01:23	

Data Download Date/Time: 12/16/2017 08:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001263

INSTRUMENT REGISTERED TO: INDIAN HARBOUR PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELAUGHTER,MARC,C
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: DELAUGHTER,MARC
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 12/16/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:57

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 12/17/2017 03:02:46

Diagnostics Check	OK	02:56
Air Blank	0.000	02:57
Control Test	0.079	02:57
Air Blank	0.000	02:58
Subject Sample # 1	0.181	02:58
Breath Volume	1.531	02:58
Air Blank	0.000	02:58
Air Blank	0.000	03:00
Subject Sample # 2	0.173	03:01
Breath Volume	1.312	03:01
Air Blank	0.000	03:01
Control Test	0.079	03:02
Air Blank	0.000	03:02
Diagnostics Check	OK	03:02

Data Download Date/Time: 01/28/2018 01:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001263

INSTRUMENT REGISTERED TO: INDIAN HARBOUR PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRIEBELL,DON,R
BREATH TEST OPERATOR AGENCY: Satellite Beach PD
ARRESTING OFFICER: TRIEBELL,DON
ARRESTING OFFICER AGENCY: Satellite Beach PD
LAST AGENCY INSPECTION: 12/16/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: TRIPP

DATE/TIME OF TEST: 12/23/2017 03:19:25

Diagnostics Check	OK	03:12
Air Blank	0.000	03:12
Control Test	0.081	03:13
Air Blank	0.000	03:13
Subject Sample # 1	0.163	03:14
Breath Volume	2.453	03:14
Air Blank	0.000	03:15
Air Blank	0.000	03:17
Subject Sample # 2	0.159	03:17
Breath Volume	1.824	03:17
Air Blank	0.000	03:18
Control Test	0.080	03:18
Air Blank	0.000	03:19
Diagnostics Check	OK	03:19

Data Download Date/Time: 01/28/2018 01:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001263

INSTRUMENT REGISTERED TO: INDIAN HARBOUR PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JAQUET,RENE,
BREATH TEST OPERATOR AGENCY: Indian Harbour Beach
ARRESTING OFFICER: SWENSON,IVAN
ARRESTING OFFICER AGENCY: Indian Harbour Beach
LAST AGENCY INSPECTION: 12/16/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:08

SUBJECT'S LAST NAME: BELL

DATE/TIME OF TEST: 12/30/2017 01:33:10

Diagnostics Check	OK	01:30	
Air Blank	0.000	01:30	
Control Test	0.080	01:31	
Air Blank	0.000	01:31	
Subject Sample	# 1	01:31	Subject Test Refused
Breath Volume	0.000	01:31	
Air Blank	0.000	01:32	
Control Test	0.080	01:32	
Air Blank	0.000	01:33	
Diagnostics Check	OK	01:33	

Data Download Date/Time: 01/28/2018 01:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLEVELAND, WILLIAM, S
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: DRULARD, BRIAN
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 11/16/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:13

SUBJECT'S LAST NAME: VIRGIL

DATE/TIME OF TEST: 12/03/2017 00:39:36

Diagnostics Check	OK	00:33
Air Blank	0.000	00:33
Control Test	0.080	00:34
Air Blank	0.000	00:34
Subject Sample # 1	0.008	00:35
Breath Volume	2.183	00:35
Air Blank	0.000	00:35
Air Blank	0.000	00:37
Subject Sample # 2	0.009	00:38
Breath Volume	2.953	00:38
Air Blank	0.000	00:38
Control Test	0.081	00:39
Air Blank	0.000	00:39
Diagnostics Check	OK	00:39

Data Download Date/Time: 01/02/2018 20:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERREIRA,ZACHARY,J
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: FERREIRA,ZACHARY
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 11/16/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:16

SUBJECT'S LAST NAME: 17596460

DATE/TIME OF TEST: 12/05/2017 00:55:37

Diagnostics Check	OK	00:48
Air Blank	0.000	00:49
Control Test	0.081	00:49
Air Blank	0.000	00:49
Subject Sample # 1	0.000	00:51
Breath Volume	1.839	00:51
Air Blank	0.000	00:51
Air Blank	0.000	00:53
Subject Sample # 2	0.000	00:54
Breath Volume	1.953	00:54
Air Blank	0.000	00:54
Control Test	0.080	00:55
Air Blank	0.000	00:55
Diagnostics Check	OK	00:55

Data Download Date/Time: 01/02/2018 20:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAVN,SUSANNE,M
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: QUINTAL,VIRGEL
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 11/16/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:55

SUBJECT'S LAST NAME: GORDON

DATE/TIME OF TEST: 12/06/2017 23:24:58

Diagnostics Check	OK	23:17
Air Blank	0.000	23:17
Control Test	0.081	23:17
Air Blank	0.000	23:18
Subject Sample # 1	0.212	23:18
Breath Volume	1.714	23:18
Air Blank	0.000	23:19
Air Blank	0.000	23:21
Subject Sample # 2	0.215	23:23
Breath Volume	1.460	23:23
Air Blank	0.000	23:24
Control Test	0.078	23:24
Air Blank	0.000	23:24
Diagnostics Check	OK	23:24

Data Download Date/Time: 01/02/2018 20:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STANTON,KEVIN,J
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: KING,AMANDA
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 11/16/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 07:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:43

SUBJECT'S LAST NAME: WATLOCK

DATE/TIME OF TEST: 12/13/2017 08:09:39

Diagnostics Check	OK	08:03
Air Blank	0.000	08:04
Control Test	0.081	08:04
Air Blank	0.000	08:04
Subject Sample # 1	0.000	08:05
Breath Volume	2.804	08:05
Air Blank	0.000	08:05
Air Blank	0.000	08:07
Subject Sample # 2	0.000	08:08
Breath Volume	2.851	08:08
Air Blank	0.000	08:08
Control Test	0.081	08:09
Air Blank	0.000	08:09
Diagnostics Check	OK	08:09

Data Download Date/Time: 01/02/2018 20:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOORE,KIMBERLY,N
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: WEIMER,ADAM
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 11/16/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:39

SUBJECT'S LAST NAME: LAKIN

DATE/TIME OF TEST: 12/23/2017 03:40:54

Diagnostics Check	OK	03:34
Air Blank	0.000	03:34
Control Test	0.080	03:35
Air Blank	0.000	03:35
Subject Sample # 1	0.097	03:36
Breath Volume	1.875	03:36
Air Blank	0.000	03:36
Air Blank	0.000	03:38
Subject Sample # 2	0.099	03:39
Breath Volume	1.652	03:39
Air Blank	0.000	03:39
Control Test	0.081	03:40
Air Blank	0.000	03:40
Diagnostics Check	OK	03:40

Data Download Date/Time: 01/02/2018 20:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAILEY,NALICIA,K
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: BAILEY,NALICIA
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 12/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: KATIPUNAN

DATE/TIME OF TEST: 12/29/2017 04:36:31

Diagnostics Check	OK	04:30
Air Blank	0.000	04:30
Control Test	0.080	04:30
Air Blank	0.000	04:31
Subject Sample # 1	0.123	04:32
Breath Volume	3.042	04:32
Air Blank	0.000	04:32
Air Blank	0.000	04:34
Subject Sample # 2	0.127	04:35
Breath Volume	2.941	04:35
Air Blank	0.000	04:35
Control Test	0.079	04:35
Air Blank	0.000	04:36
Diagnostics Check	OK	04:36

Data Download Date/Time: 01/02/2018 20:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001266

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LUTZ,EDWIN,J

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: LUTZ,EDWIN

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 11/26/2017

Date or Time Changed: No

DRY GAS STANDARD LOT #: 22817080A4

DRY GAS EXPIRY DATE: 10/05/2019

VIOLATION CODE: DUI

ARREST TIME: 02:08

GENDER: Male

OBSERVATION PERIOD BEGAN: 01:52

SUBJECT'S LAST NAME: HAMRICK

DATE/TIME OF TEST: 12/12/2017 02:18:24

Diagnostics Check	OK	02:12
Air Blank	0.000	02:12
Control Test	0.081	02:13
Air Blank	0.000	02:13
Subject Sample # 1	0.148	02:14
Breath Volume	2.378	02:14
Air Blank	0.000	02:14
Air Blank	0.000	02:16
Subject Sample # 2	0.144	02:16
Breath Volume	2.449	02:16
Air Blank	0.000	02:17
Control Test	0.080	02:17
Air Blank	0.000	02:18
Diagnostics Check	OK	02:18

Data Download Date/Time: 12/26/2017 18:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001266

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMILTON,J,W
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: HAMILTON,J
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 11/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: FLANNERY

DATE/TIME OF TEST: 12/13/2017 02:03:54

Diagnostics Check		OK	01:48	
Air Blank		0.000	01:48	
Control Test		0.080	01:49	
Air Blank		0.000	01:49	
Subject Sample	# 1	0.193	01:50	
Breath Volume		1.300	01:50	
Air Blank		0.000	01:51	
Air Blank		0.000	01:53	
Subject Sample	# 2	0.167	01:56	Volume Not Met
Breath Volume		0.199	01:56	
Air Blank		0.000	01:57	
Air Blank		0.000	01:59	
Subject Sample	# 3	0.168	02:02	Volume Not Met
Breath Volume		0.761	02:02	
Air Blank		0.000	02:02	
Control Test		0.079	02:03	
Air Blank		0.000	02:03	
Diagnostics Check		OK	02:03	

Data Download Date/Time: 12/26/2017 18:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001266

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUTZ,EDWIN,J
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: HAMILTON,JOSEPH
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 11/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: FLANNERY

DATE/TIME OF TEST: 12/13/2017 02:12:14

Diagnostics Check		OK	02:06
Air Blank		0.000	02:06
Control Test		0.078	02:06
Air Blank		0.000	02:07
Subject Sample	# 1	0.214	02:07
Breath Volume		1.382	02:07
Air Blank		0.000	02:08
Air Blank		0.000	02:10
Subject Sample	# 2	0.202	02:10
Breath Volume		1.371	02:10
Air Blank		0.000	02:11
Control Test		0.078	02:11
Air Blank		0.000	02:12
Diagnostics Check		OK	02:12

Data Download Date/Time: 12/26/2017 18:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001266

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUTZ,EDWIN,J
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: LUTZ,EDWIN
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 11/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 12/13/2017 23:01:52

Diagnostics Check	OK	22:55
Air Blank	0.000	22:55
Control Test	0.080	22:56
Air Blank	0.000	22:56
Subject Sample # 1	0.138	22:57
Breath Volume	2.746	22:57
Air Blank	0.000	22:57
Air Blank	0.000	22:59
Subject Sample # 2	0.143	23:00
Breath Volume	2.652	23:00
Air Blank	0.000	23:00
Control Test	0.079	23:01
Air Blank	0.000	23:01
Diagnostics Check	OK	23:01

Data Download Date/Time: 12/26/2017 18:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001266

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUTZ,EDWIN,J
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: LUTZ,EDWIN
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 11/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: PLUMMER

DATE/TIME OF TEST: 12/19/2017 23:37:25

Diagnostics Check	OK	23:31
Air Blank	0.000	23:31
Control Test	0.079	23:31
Air Blank	0.000	23:32
Subject Sample # 1	0.044	23:33
Breath Volume	2.566	23:33
Air Blank	0.000	23:33
Air Blank	0.000	23:35
Subject Sample # 2	0.045	23:35
Breath Volume	2.578	23:35
Air Blank	0.000	23:36
Control Test	0.080	23:36
Air Blank	0.000	23:37
Diagnostics Check	OK	23:37

Data Download Date/Time: 12/26/2017 18:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001266

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUTZ,EDWIN,J
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: LUTZ,EDWIN
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 11/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:04

SUBJECT'S LAST NAME: PEREIRA

DATE/TIME OF TEST: 12/24/2017 02:35:16

Diagnostics Check		OK	02:28
Air Blank		0.000	02:28
Control Test		0.079	02:29
Air Blank		0.000	02:29
Subject Sample	# 1	0.175	02:30
Breath Volume		1.464	02:30
Air Blank		0.000	02:30
Air Blank		0.000	02:32
Subject Sample	# 2	0.169	02:33
Breath Volume		1.445	02:33
Air Blank		0.000	02:34
Control Test		0.079	02:34
Air Blank		0.000	02:35
Diagnostics Check		OK	02:35

Data Download Date/Time: 12/26/2017 18:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001266

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUTZ,EDWIN,J
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: LUTZ,EDWIN
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 11/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: LIZOTTE

DATE/TIME OF TEST: 12/24/2017 03:35:09

Diagnostics Check		OK	03:28
Air Blank		0.000	03:29
Control Test		0.080	03:29
Air Blank		0.000	03:29
Subject Sample	# 1	0.077	03:30
Breath Volume		3.074	03:30
Air Blank		0.000	03:31
Air Blank		0.000	03:33
Subject Sample	# 2	0.075	03:33
Breath Volume		2.941	03:33
Air Blank		0.000	03:34
Control Test		0.079	03:34
Air Blank		0.000	03:35
Diagnostics Check		OK	03:35

Data Download Date/Time: 12/26/2017 18:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001266

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LUTZ,EDWIN,J

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: LUTZ,EDWIN

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 11/26/2017

Date or Time Changed: No

DRY GAS STANDARD LOT #: 22817080A4

DRY GAS EXPIRY DATE: 10/05/2019

VIOLATION CODE: DUI

ARREST TIME: 20:34

GENDER: Male

OBSERVATION PERIOD BEGAN: 20:37

SUBJECT'S LAST NAME: BISHOP

DATE/TIME OF TEST: 12/24/2017 21:03:33

Diagnostics Check	OK	20:57
Air Blank	0.000	20:58
Control Test	0.079	20:58
Air Blank	0.000	20:58
Subject Sample # 1	0.182	20:59
Breath Volume	2.117	20:59
Air Blank	0.000	20:59
Air Blank	0.000	21:01
Subject Sample # 2	0.191	21:02
Breath Volume	2.000	21:02
Air Blank	0.000	21:02
Control Test	0.079	21:02
Air Blank	0.000	21:03
Diagnostics Check	OK	21:03

Data Download Date/Time: 12/26/2017 18:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001266

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUTZ,EDWIN,J
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: LUTZ,EDWIN
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 12/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:54

SUBJECT'S LAST NAME: PLUMMER

DATE/TIME OF TEST: 12/29/2017 23:20:47

Diagnostics Check	OK	23:14
Air Blank	0.000	23:15
Control Test	0.079	23:15
Air Blank	0.000	23:16
Subject Sample # 1	0.087	23:16
Breath Volume	2.382	23:16
Air Blank	0.000	23:17
Air Blank	0.000	23:18
Subject Sample # 2	0.089	23:19
Breath Volume	2.539	23:19
Air Blank	0.000	23:19
Control Test	0.078	23:20
Air Blank	0.000	23:20
Diagnostics Check	OK	23:20

Data Download Date/Time: 01/30/2018 14:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001266

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LUTZ,EDWIN,J

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: LUTZ,EDWIN

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 12/26/2017

Date or Time Changed: No

DRY GAS STANDARD LOT #: 22817080A4

DRY GAS EXPIRY DATE: 10/05/2019

VIOLATION CODE: DUI

ARREST TIME: 23:14

GENDER: Male

OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: VNUK

DATE/TIME OF TEST: 12/30/2017 23:23:33

Diagnostics Check	OK	23:20	
Air Blank	0.000	23:20	
Control Test	0.080	23:21	
Air Blank	0.000	23:21	
Subject Sample	# 1	23:22	Subject Test Refused
Breath Volume	0.000	23:22	
Air Blank	0.000	23:22	
Control Test	0.081	23:22	
Air Blank	0.000	23:23	
Diagnostics Check	OK	23:23	

Data Download Date/Time: 01/30/2018 14:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANCION,CALIX,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: ANCION,CALIX
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 11/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 06:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:28

SUBJECT'S LAST NAME: JAUME

DATE/TIME OF TEST: 12/03/2017 07:55:42

Diagnostics Check	OK	07:48
Air Blank	0.000	07:48
Control Test	0.080	07:49
Air Blank	0.000	07:49
Subject Sample # 1	0.112	07:50
Breath Volume	1.546	07:50
Air Blank	0.000	07:51
Air Blank	0.000	07:53
Subject Sample # 2	0.111	07:54
Breath Volume	1.500	07:54
Air Blank	0.000	07:54
Control Test	0.079	07:55
Air Blank	0.000	07:55
Diagnostics Check	OK	07:55

Data Download Date/Time: 01/02/2018 23:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: WILKINS,AARON
ARRESTING OFFICER AGENCY: Winter Springs PD
LAST AGENCY INSPECTION: 11/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 23:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 12/16/2017 00:51:36

Diagnostics Check	OK	00:45
Air Blank	0.000	00:45
Control Test	0.080	00:46
Air Blank	0.000	00:46
Subject Sample # 1	0.243	00:47
Breath Volume	4.117	00:47
Air Blank	0.000	00:47
Air Blank	0.000	00:49
Subject Sample # 2	0.247	00:50
Breath Volume	3.285	00:50
Air Blank	0.000	00:50
Control Test	0.080	00:51
Air Blank	0.000	00:51
Diagnostics Check	OK	00:51

Data Download Date/Time: 01/02/2018 23:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: DELLOVA,JOSEPH
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 11/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:06

SUBJECT'S LAST NAME: VAN AUKEN

DATE/TIME OF TEST: 12/16/2017 02:33:19

Diagnostics Check	OK	02:26
Air Blank	0.000	02:27
Control Test	0.080	02:27
Air Blank	0.000	02:27
Subject Sample # 1	0.222	02:28
Breath Volume	1.371	02:28
Air Blank	0.000	02:29
Air Blank	0.000	02:30
Subject Sample # 2	0.221	02:31
Breath Volume	1.332	02:31
Air Blank	0.000	02:32
Control Test	0.080	02:32
Air Blank	0.000	02:33
Diagnostics Check	OK	02:33

Data Download Date/Time: 01/02/2018 23:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA, RAY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: ANCION, CALIX
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 11/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: FERNANDEZ

DATE/TIME OF TEST: 12/17/2017 03:22:00

Diagnostics Check	OK	03:15
Air Blank	0.000	03:16
Control Test	0.081	03:16
Air Blank	0.000	03:16
Subject Sample # 1	0.052	03:17
Breath Volume	3.414	03:17
Air Blank	0.000	03:18
Air Blank	0.000	03:20
Subject Sample # 2	0.044	03:20
Breath Volume	1.386	03:20
Air Blank	0.000	03:21
Control Test	0.081	03:21
Air Blank	0.000	03:21
Diagnostics Check	OK	03:22

Data Download Date/Time: 01/02/2018 23:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA, RAY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: STUBBS, MATHEW
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 11/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 04:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: SANTIAGO

DATE/TIME OF TEST: 12/21/2017 05:10:03

Diagnostics Check	OK	05:03
Air Blank	0.000	05:04
Control Test	0.080	05:04
Air Blank	0.000	05:05
Subject Sample # 1	0.178	05:05
Breath Volume	3.253	05:05
Air Blank	0.000	05:06
Air Blank	0.000	05:08
Subject Sample # 2	0.187	05:08
Breath Volume	2.406	05:08
Air Blank	0.000	05:09
Control Test	0.080	05:09
Air Blank	0.000	05:09
Diagnostics Check	OK	05:10

Data Download Date/Time: 01/02/2018 23:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: MEADOWS,SHAWN
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 11/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 00:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: GIGLIO

DATE/TIME OF TEST: 12/23/2017 01:22:06

Diagnostics Check	OK	01:15
Air Blank	0.000	01:16
Control Test	0.080	01:16
Air Blank	0.000	01:17
Subject Sample # 1	0.088	01:17
Breath Volume	3.253	01:17
Air Blank	0.000	01:18
Air Blank	0.000	01:20
Subject Sample # 2	0.088	01:20
Breath Volume	2.917	01:20
Air Blank	0.000	01:21
Control Test	0.081	01:21
Air Blank	0.000	01:22
Diagnostics Check	OK	01:22

Data Download Date/Time: 01/02/2018 23:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: HALL,JONATHAN
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 11/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 03:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:24

SUBJECT'S LAST NAME: MCGARRY

DATE/TIME OF TEST: 12/23/2017 03:50:39

Diagnostics Check		OK	03:44
Air Blank		0.000	03:44
Control Test		0.081	03:45
Air Blank		0.000	03:45
Subject Sample	# 1	0.233	03:46
Breath Volume		3.695	03:46
Air Blank		0.000	03:46
Air Blank		0.000	03:48
Subject Sample	# 2	0.228	03:49
Breath Volume		3.703	03:49
Air Blank		0.000	03:49
Control Test		0.080	03:50
Air Blank		0.000	03:50
Diagnostics Check		OK	03:50

Data Download Date/Time: 01/02/2018 23:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: MEADOWS,SHAWN
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 11/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 03:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: RIOS

DATE/TIME OF TEST: 12/23/2017 04:22:37

Diagnostics Check	OK	04:15
Air Blank	0.000	04:15
Control Test	0.080	04:16
Air Blank	0.000	04:16
Subject Sample # 1	0.158	04:17
Breath Volume	3.105	04:17
Air Blank	0.000	04:18
Air Blank	0.000	04:20
Subject Sample # 2	0.149	04:21
Breath Volume	2.320	04:21
Air Blank	0.000	04:21
Control Test	0.080	04:22
Air Blank	0.000	04:22
Diagnostics Check	OK	04:22

Data Download Date/Time: 01/02/2018 23:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEADOWS,SHAWN,M
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: VANHEUSEN,ROBERT
ARRESTING OFFICER AGENCY: Oviedo PD
LAST AGENCY INSPECTION: 11/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 16:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:31

SUBJECT'S LAST NAME: WUNDER

DATE/TIME OF TEST: 12/23/2017 18:09:46

Diagnostics Check	OK	17:58
Air Blank	0.000	17:59
Control Test	0.079	17:59
Air Blank	0.000	18:00
Subject Sample # 1	0.130	18:03
Breath Volume	1.453	18:03
Air Blank	0.000	18:03
Air Blank	0.000	18:05
Subject Sample # 2	0.138	18:08
Breath Volume	1.304	18:08
Air Blank	0.000	18:08
Control Test	0.078	18:09
Air Blank	0.000	18:09
Diagnostics Check	OK	18:09

Data Download Date/Time: 01/02/2018 23:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA, RAY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: MOSELEY, SCOTT
ARRESTING OFFICER AGENCY: Oviedo PD
LAST AGENCY INSPECTION: 11/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 02:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: KELLY

DATE/TIME OF TEST: 12/26/2017 03:23:06

Diagnostics Check	OK	03:14
Air Blank	0.000	03:14
Control Test	0.081	03:15
Air Blank	0.000	03:15
Subject Sample # 1	0.172	03:17
Breath Volume	2.191	03:17
Air Blank	0.000	03:18
Air Blank	0.000	03:20
Subject Sample # 2	0.177	03:21
Breath Volume	2.542	03:21
Air Blank	0.000	03:22
Control Test	0.080	03:22
Air Blank	0.000	03:23
Diagnostics Check	OK	03:23

Data Download Date/Time: 01/02/2018 23:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA, RAY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: TEEHAN, MATTHEW
ARRESTING OFFICER AGENCY: Altamonte Springs PD
LAST AGENCY INSPECTION: 11/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 04:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:50

SUBJECT'S LAST NAME: TRAN

DATE/TIME OF TEST: 12/03/2017 05:17:34

Diagnostics Check	OK	05:11
Air Blank	0.000	05:11
Control Test	0.081	05:12
Air Blank	0.000	05:12
Subject Sample # 1	0.109	05:13
Breath Volume	1.960	05:13
Air Blank	0.000	05:13
Air Blank	0.000	05:15
Subject Sample # 2	0.105	05:16
Breath Volume	1.296	05:16
Air Blank	0.000	05:16
Control Test	0.080	05:16
Air Blank	0.000	05:17
Diagnostics Check	OK	05:17

Data Download Date/Time: 01/02/2018 23:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA, RAY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: RITTER, ROBERT
ARRESTING OFFICER AGENCY: Lake Mary PD
LAST AGENCY INSPECTION: 11/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 05:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:45

SUBJECT'S LAST NAME: CHAVEZ GABRIEL

DATE/TIME OF TEST: 12/03/2017 06:12:14

Diagnostics Check	OK	06:05
Air Blank	0.000	06:06
Control Test	0.080	06:06
Air Blank	0.000	06:07
Subject Sample # 1	0.139	06:07
Breath Volume	2.527	06:07
Air Blank	0.000	06:08
Air Blank	0.000	06:10
Subject Sample # 2	0.131	06:10
Breath Volume	2.769	06:10
Air Blank	0.000	06:11
Control Test	0.081	06:11
Air Blank	0.000	06:12
Diagnostics Check	OK	06:12

Data Download Date/Time: 01/02/2018 23:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRAWFORD,L,P
BREATH TEST OPERATOR AGENCY: Lake Mary PD
ARRESTING OFFICER: MUNOZ,G
ARRESTING OFFICER AGENCY: Sanford PD
LAST AGENCY INSPECTION: 11/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 08:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:00

SUBJECT'S LAST NAME: STEENERSON

DATE/TIME OF TEST: 12/06/2017 09:29:26

Diagnostics Check	OK	09:22
Air Blank	0.000	09:23
Control Test	0.080	09:23
Air Blank	0.000	09:24
Subject Sample # 1	0.254	09:24
Breath Volume	3.839	09:24
Air Blank	0.000	09:25
Air Blank	0.000	09:27
Subject Sample # 2	0.251	09:27
Breath Volume	3.882	09:27
Air Blank	0.000	09:28
Control Test	0.081	09:28
Air Blank	0.000	09:29
Diagnostics Check	OK	09:29

Data Download Date/Time: 01/02/2018 23:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA, RAY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: KINKOPF, TIMOTHY
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 11/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 20:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:40

SUBJECT'S LAST NAME: COPELAND

DATE/TIME OF TEST: 12/06/2017 21:17:32

Diagnostics Check	OK	21:11
Air Blank	0.000	21:11
Control Test	0.081	21:12
Air Blank	0.000	21:12
Subject Sample # 1	0.165	21:13
Breath Volume	2.226	21:13
Air Blank	0.000	21:13
Air Blank	0.000	21:15
Subject Sample # 2	0.164	21:15
Breath Volume	2.179	21:15
Air Blank	0.000	21:16
Control Test	0.081	21:16
Air Blank	0.000	21:17
Diagnostics Check	OK	21:17

Data Download Date/Time: 01/02/2018 23:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: MEADOWS,SHAWN
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 11/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 01:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: TODD

DATE/TIME OF TEST: 12/09/2017 02:54:56

Diagnostics Check	OK	02:46
Air Blank	0.000	02:46
Control Test	0.082	02:47
Air Blank	0.000	02:47
Subject Sample # 1	0.196	02:50
Breath Volume	1.687	02:50
Air Blank	0.000	02:50
Air Blank	0.000	02:52
Subject Sample # 2	0.195	02:53
Breath Volume	1.859	02:53
Air Blank	0.000	02:54
Control Test	0.082	02:54
Air Blank	0.000	02:54
Diagnostics Check	OK	02:54

Data Download Date/Time: 01/02/2018 23:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: GRAFFY,THOMAS
ARRESTING OFFICER AGENCY: Casselberry PD
LAST AGENCY INSPECTION: 11/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 02:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: BOX

DATE/TIME OF TEST: 12/09/2017 03:09:06

Diagnostics Check	OK	03:02
Air Blank	0.000	03:03
Control Test	0.080	03:03
Air Blank	0.000	03:04
Subject Sample # 1	0.150	03:04
Breath Volume	5.187	03:04
Air Blank	0.000	03:05
Air Blank	0.000	03:07
Subject Sample # 2	0.147	03:07
Breath Volume	4.910	03:07
Air Blank	0.000	03:08
Control Test	0.081	03:08
Air Blank	0.000	03:09
Diagnostics Check	OK	03:09

Data Download Date/Time: 01/02/2018 23:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: MOREL,ANTHONY
ARRESTING OFFICER AGENCY: Winter Springs PD
LAST AGENCY INSPECTION: 11/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:03

SUBJECT'S LAST NAME: CASTRO

DATE/TIME OF TEST: 12/10/2017 02:29:51

Diagnostics Check	OK	02:23
Air Blank	0.000	02:24
Control Test	0.080	02:24
Air Blank	0.000	02:25
Subject Sample # 1	0.000	02:25
Breath Volume	1.921	02:25
Air Blank	0.000	02:26
Air Blank	0.000	02:28
Subject Sample # 2	0.000	02:28
Breath Volume	3.281	02:28
Air Blank	0.000	02:28
Control Test	0.078	02:29
Air Blank	0.000	02:29
Diagnostics Check	OK	02:29

Data Download Date/Time: 01/02/2018 23:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: VANHEUSEN,ROBERT
ARRESTING OFFICER AGENCY: Oviedo PD
LAST AGENCY INSPECTION: 11/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:32

SUBJECT'S LAST NAME: POPE

DATE/TIME OF TEST: 12/10/2017 03:01:13

Diagnostics Check	OK	02:53
Air Blank	0.000	02:53
Control Test	0.082	02:54
Air Blank	0.000	02:54
Subject Sample # 1	0.106	02:55
Breath Volume	2.726	02:55
Air Blank	0.000	02:56
Air Blank	0.000	02:58
Subject Sample # 2	0.113	02:59
Breath Volume	3.574	02:59
Air Blank	0.000	03:00
Control Test	0.081	03:00
Air Blank	0.000	03:01
Diagnostics Check	OK	03:01

Data Download Date/Time: 01/02/2018 23:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: MORRIS,ERIC
ARRESTING OFFICER AGENCY: Winter Springs PD
LAST AGENCY INSPECTION: 11/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 03:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:01

SUBJECT'S LAST NAME: BARKER

DATE/TIME OF TEST: 12/10/2017 04:28:06

Diagnostics Check	OK	04:21
Air Blank	0.000	04:22
Control Test	0.082	04:22
Air Blank	0.000	04:23
Subject Sample # 1	0.175	04:23
Breath Volume	3.050	04:23
Air Blank	0.000	04:24
Air Blank	0.000	04:26
Subject Sample # 2	0.169	04:26
Breath Volume	2.886	04:26
Air Blank	0.000	04:27
Control Test	0.081	04:27
Air Blank	0.000	04:28
Diagnostics Check	OK	04:28

Data Download Date/Time: 01/02/2018 23:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: REHDER,PATRICK
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 11/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 04:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:04

SUBJECT'S LAST NAME: COLLAZO

DATE/TIME OF TEST: 12/10/2017 05:40:45

Diagnostics Check	OK	05:34
Air Blank	0.000	05:34
Control Test	0.081	05:35
Air Blank	0.000	05:35
Subject Sample # 1	0.152	05:36
Breath Volume	2.453	05:36
Air Blank	0.000	05:36
Air Blank	0.000	05:38
Subject Sample # 2	0.154	05:39
Breath Volume	2.644	05:39
Air Blank	0.000	05:39
Control Test	0.081	05:40
Air Blank	0.000	05:40
Diagnostics Check	OK	05:40

Data Download Date/Time: 01/02/2018 23:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: VANHEUSEN,ROBERT
ARRESTING OFFICER AGENCY: Oviedo PD
LAST AGENCY INSPECTION: 11/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 00:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:24

SUBJECT'S LAST NAME: MARTIN

DATE/TIME OF TEST: 12/15/2017 02:01:09

Diagnostics Check	OK	01:54
Air Blank	0.000	01:54
Control Test	0.082	01:55
Air Blank	0.000	01:55
Subject Sample # 1	0.194	01:56
Breath Volume	1.265	01:56
Air Blank	0.000	01:57
Air Blank	0.000	01:58
Subject Sample # 2	0.209	01:59
Breath Volume	1.214	01:59
Air Blank	0.000	02:00
Control Test	0.082	02:00
Air Blank	0.000	02:01
Diagnostics Check	OK	02:01

Data Download Date/Time: 01/02/2018 23:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: CASSATO,KAYLA
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 11/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: MOSES

DATE/TIME OF TEST: 12/15/2017 03:16:54

Diagnostics Check	OK	03:10
Air Blank	0.000	03:11
Control Test	0.081	03:11
Air Blank	0.000	03:11
Subject Sample # 1	0.213	03:12
Breath Volume	2.832	03:12
Air Blank	0.000	03:13
Air Blank	0.000	03:14
Subject Sample # 2	0.211	03:15
Breath Volume	2.664	03:15
Air Blank	0.000	03:15
Control Test	0.081	03:16
Air Blank	0.000	03:16
Diagnostics Check	OK	03:16

Data Download Date/Time: 01/02/2018 23:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: ANCION,CALIX
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 11/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 03:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:38

SUBJECT'S LAST NAME: OLIVER

DATE/TIME OF TEST: 12/15/2017 05:07:18

Diagnostics Check	OK	04:59
Air Blank	0.000	05:00
Control Test	0.081	05:00
Air Blank	0.000	05:01
Subject Sample # 1	0.084	05:02
Breath Volume	1.718	05:02
Air Blank	0.000	05:03
Air Blank	0.000	05:05
Subject Sample # 2	0.082	05:05
Breath Volume	1.976	05:05
Air Blank	0.000	05:06
Control Test	0.081	05:06
Air Blank	0.000	05:07
Diagnostics Check	OK	05:07

Data Download Date/Time: 01/02/2018 23:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES CRUZ,KENNY,Y
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: STEWART,TIMOTHY
ARRESTING OFFICER AGENCY: Casselberry PD
LAST AGENCY INSPECTION: 11/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 06:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:03

SUBJECT'S LAST NAME: MCGAHA

DATE/TIME OF TEST: 12/23/2017 07:37:43

Diagnostics Check		OK	07:26	
Air Blank		0.000	07:26	
Control Test		0.081	07:26	
Air Blank		0.000	07:27	
Subject Sample	# 1	0.147	07:30	Volume Not Met
Breath Volume		0.234	07:30	
Air Blank		0.000	07:31	
Air Blank		0.000	07:33	
Subject Sample	# 2	0.133	07:36	Volume Not Met
Breath Volume		0.121	07:36	
Air Blank		0.000	07:36	
Control Test		0.078	07:37	
Air Blank		0.000	07:37	
Diagnostics Check		OK	07:37	

Data Download Date/Time: 01/02/2018 23:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANCION,CALIX,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: ANCION,CALIX
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 11/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 21:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: CORONA

DATE/TIME OF TEST: 12/23/2017 22:16:00

Diagnostics Check	OK	22:09
Air Blank	0.000	22:10
Control Test	0.080	22:10
Air Blank	0.000	22:10
Subject Sample # 1	0.292	22:11
Breath Volume	2.890	22:11
Air Blank	0.000	22:12
Air Blank	0.000	22:13
Subject Sample # 2	0.298	22:14
Breath Volume	2.539	22:14
Air Blank	0.000	22:15
Control Test	0.079	22:15
Air Blank	0.000	22:15
Diagnostics Check	OK	22:16

Data Download Date/Time: 01/02/2018 23:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: MORRIS,ERIC
ARRESTING OFFICER AGENCY: Winter Springs PD
LAST AGENCY INSPECTION: 11/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 00:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:37

SUBJECT'S LAST NAME: MCKOWN

DATE/TIME OF TEST: 12/24/2017 01:09:54

Diagnostics Check	OK	01:03
Air Blank	0.000	01:04
Control Test	0.079	01:04
Air Blank	0.000	01:04
Subject Sample # 1	0.144	01:05
Breath Volume	1.746	01:05
Air Blank	0.000	01:05
Air Blank	0.000	01:07
Subject Sample # 2	0.156	01:08
Breath Volume	2.437	01:08
Air Blank	0.000	01:09
Control Test	0.080	01:09
Air Blank	0.000	01:09
Diagnostics Check	OK	01:09

Data Download Date/Time: 01/02/2018 23:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANCION,CALIX,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: ANCION,CALIX'
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 11/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 15:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:20

SUBJECT'S LAST NAME: CARRANZA

DATE/TIME OF TEST: 12/28/2017 19:48:05

Diagnostics Check	OK	19:40
Air Blank	0.000	19:41
Control Test	0.080	19:41
Air Blank	0.000	19:42
Subject Sample # 1	0.218	19:42
Breath Volume	1.597	19:42
Air Blank	0.000	19:43
Air Blank	0.000	19:45
Subject Sample # 2	0.216	19:46
Breath Volume	1.542	19:46
Air Blank	0.000	19:47
Control Test	0.080	19:47
Air Blank	0.000	19:48
Diagnostics Check	OK	19:48

Data Download Date/Time: 01/02/2018 23:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: ARGENTINE,BRITTANY
ARRESTING OFFICER AGENCY: Oviedo PD
LAST AGENCY INSPECTION: 11/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 21:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:13

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 12/29/2017 22:41:23

Diagnostics Check	OK	22:33
Air Blank	0.000	22:33
Control Test	0.080	22:34
Air Blank	0.000	22:34
Subject Sample # 1	0.157	22:36
Breath Volume	1.171	22:36
Air Blank	0.000	22:36
Air Blank	0.000	22:38
Subject Sample # 2	0.167	22:39
Breath Volume	1.652	22:39
Air Blank	0.000	22:40
Control Test	0.081	22:40
Air Blank	0.000	22:41
Diagnostics Check	OK	22:41

Data Download Date/Time: 01/02/2018 23:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: MEADOWS,SHAWN
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 11/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: BOWIE

DATE/TIME OF TEST: 12/30/2017 01:56:59

Diagnostics Check	OK	01:50
Air Blank	0.000	01:51
Control Test	0.080	01:51
Air Blank	0.000	01:51
Subject Sample # 1	0.091	01:52
Breath Volume	4.851	01:52
Air Blank	0.000	01:53
Air Blank	0.000	01:54
Subject Sample # 2	0.087	01:55
Breath Volume	4.265	01:55
Air Blank	0.000	01:56
Control Test	0.080	01:56
Air Blank	0.000	01:56
Diagnostics Check	OK	01:56

Data Download Date/Time: 01/02/2018 23:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: COLBURN,MEGAN
ARRESTING OFFICER AGENCY: Casselberry PD
LAST AGENCY INSPECTION: 11/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 02:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: ROSA-REYES

DATE/TIME OF TEST: 12/30/2017 03:58:01

Diagnostics Check	OK	03:51
Air Blank	0.000	03:52
Control Test	0.081	03:52
Air Blank	0.000	03:52
Subject Sample # 1	0.157	03:53
Breath Volume	2.507	03:53
Air Blank	0.000	03:54
Air Blank	0.000	03:55
Subject Sample # 2	0.159	03:56
Breath Volume	2.015	03:56
Air Blank	0.000	03:57
Control Test	0.081	03:57
Air Blank	0.000	03:57
Diagnostics Check	OK	03:58

Data Download Date/Time: 01/02/2018 23:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANTHONY, MICHAEL,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ELLIS, ALEXANDER
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 00:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:36

SUBJECT'S LAST NAME: MAYS

DATE/TIME OF TEST: 12/22/2017 01:04:34

Diagnostics Check		OK	00:57
Air Blank		0.000	00:58
Control Test		0.081	00:58
Air Blank		0.000	00:59
Subject Sample	# 1	0.149	00:59
Breath Volume		1.832	00:59
Air Blank		0.000	01:00
Air Blank		0.000	01:02
Subject Sample	# 2	0.137	01:02
Breath Volume		1.949	01:02
Air Blank		0.000	01:03
Control Test		0.081	01:03
Air Blank		0.000	01:04
Diagnostics Check		OK	01:04

Data Download Date/Time: 12/27/2017 22:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANTHONY, MICHAEL,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KASELAK, STEPHEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 01:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:23

SUBJECT'S LAST NAME: PATEL

DATE/TIME OF TEST: 12/22/2017 01:51:23

Diagnostics Check	OK	01:44
Air Blank	0.000	01:45
Control Test	0.082	01:45
Air Blank	0.000	01:45
Subject Sample # 1	0.192	01:46
Breath Volume	1.503	01:46
Air Blank	0.000	01:47
Air Blank	0.000	01:48
Subject Sample # 2	0.197	01:49
Breath Volume	1.367	01:49
Air Blank	0.000	01:50
Control Test	0.081	01:50
Air Blank	0.000	01:51
Diagnostics Check	OK	01:51

Data Download Date/Time: 12/27/2017 22:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANTHONY, MICHAEL,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WALDO, AARON
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 11/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 22:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 12/22/2017 23:09:36

Diagnostics Check	OK	23:00	
Air Blank	0.000	23:01	
Control Test	0.082	23:01	
Air Blank	0.000	23:02	
Subject Sample	# 1	23:05	No Sample Provided
Breath Volume	0.000	23:05	
Air Blank	0.000	23:05	
Air Blank	0.000	23:07	
Subject Sample	# 2	23:08	Subject Test Refused
Breath Volume	0.000	23:08	
Air Blank	0.000	23:08	
Control Test	0.081	23:08	
Air Blank	0.000	23:09	
Diagnostics Check	OK	23:09	

Data Download Date/Time: 12/27/2017 22:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANTHONY, MICHAEL,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MOWATT, N
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 23:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:42

SUBJECT'S LAST NAME: ELLINGTON

DATE/TIME OF TEST: 12/23/2017 00:09:52

Diagnostics Check	OK	00:03
Air Blank	0.000	00:03
Control Test	0.081	00:04
Air Blank	0.000	00:04
Subject Sample # 1	0.118	00:05
Breath Volume	1.902	00:05
Air Blank	0.000	00:05
Air Blank	0.000	00:07
Subject Sample # 2	0.116	00:08
Breath Volume	1.933	00:08
Air Blank	0.000	00:08
Control Test	0.081	00:09
Air Blank	0.000	00:09
Diagnostics Check	OK	00:09

Data Download Date/Time: 12/27/2017 22:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANTHONY, MICHAEL,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BRUNER, SEAN
ARRESTING OFFICER AGENCY: Kenneth City PD
LAST AGENCY INSPECTION: 11/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: HAMPTON

DATE/TIME OF TEST: 12/23/2017 00:58:06

Diagnostics Check	OK	00:51
Air Blank	0.000	00:51
Control Test	0.082	00:51
Air Blank	0.000	00:52
Subject Sample # 1	0.007	00:52
Breath Volume	1.347	00:52
Air Blank	0.000	00:53
Air Blank	0.000	00:55
Subject Sample # 2	0.000	00:56
Breath Volume	2.019	00:56
Air Blank	0.000	00:57
Control Test	0.081	00:57
Air Blank	0.000	00:57
Diagnostics Check	OK	00:58

Data Download Date/Time: 12/27/2017 22:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANTHONY, MICHAEL,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: SILVERSTEIN, C
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 11/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 01:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: GAUSS

DATE/TIME OF TEST: 12/23/2017 02:30:12

Diagnostics Check	OK	02:23
Air Blank	0.000	02:24
Control Test	0.081	02:24
Air Blank	0.000	02:24
Subject Sample # 1	0.217	02:25
Breath Volume	3.621	02:25
Air Blank	0.000	02:26
Air Blank	0.000	02:28
Subject Sample # 2	0.217	02:28
Breath Volume	2.445	02:28
Air Blank	0.000	02:29
Control Test	0.081	02:29
Air Blank	0.000	02:30
Diagnostics Check	OK	02:30

Data Download Date/Time: 12/27/2017 22:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LANEY,DAMON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:36

SUBJECT'S LAST NAME: DEVENNEY

DATE/TIME OF TEST: 12/23/2017 00:03:21

Diagnostics Check	OK	00:00	
Air Blank	0.000	00:00	
Control Test	0.079	00:01	
Air Blank	0.000	00:01	
Subject Sample	# 1	00:01	Subject Test Refused
Breath Volume	0.000	00:01	
Air Blank	0.000	00:02	
Control Test	0.078	00:02	
Air Blank	0.000	00:03	
Diagnostics Check	OK	00:03	

Data Download Date/Time: 12/27/2017 21:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KASELAK,STEPHEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 01:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:28

SUBJECT'S LAST NAME: GOOCH

DATE/TIME OF TEST: 12/23/2017 01:55:09

Diagnostics Check	OK	01:48
Air Blank	0.000	01:49
Control Test	0.079	01:49
Air Blank	0.000	01:50
Subject Sample # 1	0.212	01:50
Breath Volume	3.425	01:50
Air Blank	0.000	01:51
Air Blank	0.000	01:53
Subject Sample # 2	0.215	01:53
Breath Volume	3.515	01:53
Air Blank	0.000	01:54
Control Test	0.077	01:54
Air Blank	0.000	01:55
Diagnostics Check	OK	01:55

Data Download Date/Time: 12/27/2017 21:09